



AAFCO Quality Assurance/Quality Control Guidelines for Feed Laboratories, 2014 Edition

Supplement to ISO/IEC 17025:2005

General Requirements for the Competence of Testing and Calibration Laboratories

ERRATA SHEET – February 2016

The following correction has been made to **VOLUME III Table 2 for Equipment, Calibration, Verification and Maintenance**

		Verification		Maintenance	
Equipment	Calibration	Parameter checked	Minimum Frequency	Recommended	Minimum Frequency
Water Purification System		Resistivity/conductivity	Weekly for Stills, Monthly for other systems, or as recommended by method	As recommended by manufacturer or per laboratory procedure	As recommended by manufacturer or per laboratory procedure