

AAFCO Check Sample Program

All Methods Report

Sort by Method

Proficiency For Individual Methods

Sample # 201344

Dried Egg Product

Pet Food Add-on



AAFCO

Check Sample Program



Issue Date : 01/31/2014

Method Code	Total # Labs Submitting	# Labs Included in Calculations	Mean	SD	Assigned Value - Robust Mean	AAFCO CS fip - Robust sd	Uncertainty (U) - Robust	% RSD - Robust	Between Labs sL	Within Labs sI	Reproducibility sR	Between Labs %RSD	Within Labs %s/d	Reproducibility %RSD	sR/sr	Average Range (R-bar)	Horwitz %RSD
001.00	9	6	4.3383	0.19916	4.3063	0.16011	0.04622	3.72%	0.16875	0.14961	0.22552	3.89%	3.45%	5.20%	1.5074	0.14333	3.21%
001.03	1		5.1000														
001.07	7	6	4.4319	0.22462	4.4319	0.25471	0.07353	5.75%	0.20603	0.12651	0.24177	4.65%	2.85%	5.46%	1.9111	0.12717	3.20%
002.00	1		58.340														
002.02	1		58.020														
002.04	2	0															
002.05	1		58.895														
002.06	111	40	59.180	0.32671	59.218	0.30350	0.03273	0.51%	0.30511	0.16522	0.34697	0.52%	0.28%	0.59%	2.1001	0.18042	2.16%
002.99	2	0															
003.00	5	4	25.058	0.59337	27.406	5.2765	1.6686	19.25%	0.55224	0.30700	0.63184	2.20%	1.23%	2.52%	2.0581	0.65200	2.43%
003.01	1		27.110														
003.06	1		24.445														
003.09	4	3	24.910	1.2315	20.255	9.3641	3.3107	46.23%	1.2157	0.27863	1.2472	4.88%	1.12%	5.01%	4.4762	0.31000	2.54%
003.10	7	4	24.230	0.35178	24.230	0.35178	0.12437	1.45%	0.34780	0.07467	0.35572	1.44%	0.31%	1.47%	4.7642	0.09000	2.48%
003.12	1		25.615														
003.13	2	0															
003.14	2	0															
003.99	5	3	27.088	3.4631	27.088	3.4631	1.4138	12.78%	3.4598	0.21545	3.4665	12.77%	0.80%	12.80%	16.090	0.20333	2.43%
004.00	6	4	0.33500	0.29852	0.69509	0.69969	0.20198	100.66%	0.29758	0.03354	0.29946	88.83%	10.01%	89.39%	8.9283	0.12833	4.22%
004.07	8	8	0.77438	0.38661	0.75420	0.40023	0.10006	53.07%	0.36287	0.18863	0.40897	46.86%	24.36%	52.81%	2.1681	0.21125	4.17%
005.00	91	25	4.3474	0.07514	4.3545	0.06496	0.00884	1.49%	0.07355	0.02173	0.07669	1.69%	0.50%	1.76%	3.5290	0.03679	3.21%
005.03	2	0															
005.05	1		4.4000														
005.99	2	0															
010.03	8	0			3.4750	1.0112	0.71500	29.10%								0.13000	3.32%
010.99	1		4.2800														
011.01	74	19	5.1006	0.21209	5.1456	0.16688	0.02639	3.24%	0.20926	0.04878	0.21487	4.10%	0.96%	4.21%	4.4050	0.06854	3.13%

Method Code	Total # Labs Submitting	# Labs Included in Calculations	Mean	SD	Assigned Value - Robust Mean	AAFCO CS ftp - Robust sd	Uncertainty (U) - Robust	% RSD - Robust	Between Labs sL	Within Labs sr	Reproducibility sR	Between Labs %RSD	Within Labs %rsd	Reproducibility %RSD	sR/sr	Average Range (R-bar)	Horwitz %RSD
011.02	2	0															
011.03	13	5	4.8300	0.32802	4.8300	0.32802	0.10373	6.79%	0.32696	0.03742	0.32909	6.77%	0.77%	6.81%	8.7953	0.04800	3.16%
012.01	1		0.08500														
013.00	6	4	30.231	4.7582	30.231	4.7582	1.6823	15.74%	4.7554	0.22976	4.7610	15.73%	0.76%	15.75%	20.722	0.24750	2.39%
013.02	79	18	29.223	1.5942	29.422	1.0907	0.17694	3.71%	1.5741	0.35680	1.6140	5.39%	1.22%	5.52%	4.5237	0.43211	2.40%
013.10	3	3	28.510	1.3599	28.510	1.3599	0.55517	4.77%	1.3559	0.14629	1.3638	4.76%	0.51%	4.78%	9.3229	0.19333	2.42%
013.12	1		31.550														
013.13	2	0															
015.41	1		135.00														
019.31	3	3	0.33038	0.03065	0.33038	0.03065	0.01251	9.28%	0.03043	0.00521	0.03087	9.21%	1.58%	9.34%	5.9281	0.00690	4.73%
019.32	1		0.38400														
019.33	1		0.37950														
019.41	9	8	0.36348	0.00983	0.36175	0.01147	0.00270	3.17%	0.00347	0.01301	0.01347	0.95%	3.58%	3.70%	1.0349	0.01243	4.66%
019.42	1		0.34550														
019.43	3	2	0.34975	0.00530	0.38067	0.05368	0.02191	14.10%	0.00500	0.00250	0.00559	1.43%	0.71%	1.60%	2.2361	0.04800	4.63%
019.44	1		0.36250														
021.41	1		0.01550														
021.43	1		0.08000														
022.32	1		1.7345														
022.41	5	4	2.0488	0.70996	3.3390	2.9499	0.93283	88.35%	0.34635	0.87645	0.94240	16.91%	42.78%	46.00%	1.0753	2.2380	13.34%
022.43	1		6.9900														
022.44	1		1.8800														
025.31	3	3	61.533	5.1501	61.533	5.1501	2.1025	8.37%	4.5358	3.4496	5.6985	7.37%	5.61%	9.26%	1.6519	4.7333	8.61%
025.32	1		70.700														
025.33	1		52.135														
025.41	9	9	60.099	5.7960	60.054	4.8800	1.1502	8.13%	5.2953	3.3328	6.2568	8.81%	5.55%	10.41%	1.8773	3.7556	8.64%
025.43	3	3	81.483	30.769	81.483	30.769	12.562	37.76%	30.192	8.3878	31.336	37.05%	10.29%	38.46%	3.7359	8.3000	8.25%
027.31	3	3	0.04422	0.00920	0.04422	0.00920	0.00376	20.81%	0.00866	0.00440	0.00971	19.58%	9.94%	21.96%	2.2085	0.00470	6.40%
027.32	1		0.04555														
027.33	1		0.04865														
027.41	8	7	0.05235	0.00155	0.05196	0.00160	0.00040	3.08%		0.00281			5.37%			0.00251	6.24%
027.43	2	0															
027.44	1		0.05035														
028.32	1		1.3585														
028.33	1		2.0000														
028.41	5	4	0.71625	0.45222	1.0730	0.88867	0.28102	82.82%	0.45104	0.04623	0.45340	62.97%	6.45%	63.30%	9.8069	0.23800	15.83%
028.44	1		0.94550														
031.01	3	3	0.57937	0.01425	0.57937	0.01425	0.00582	2.46%	0.00853	0.01615	0.01826	1.47%	2.79%	3.15%	1.1308	0.01967	4.34%

Method Code	Total # Labs Submitting	# Labs Included in Calculations	Mean	SD	Assigned Value - Robust Mean	AAFCO CS fip - Robust sd	Uncertainty (U) - Robust	% RSD - Robust	Between Labs sL	Within Labs sr	Reproducibility sR	Between Labs %RSD	Within Labs %rsd	Reproducibility %RSD	sR/sr	Average Range (R-bar)	Horwitz %RSD
031.03	1		0.65800														
031.41	9	8	0.59867	0.01739	0.60316	0.02395	0.00564	3.97%	0.01382	0.01492	0.02034	2.31%	2.49%	3.40%	1.3634	0.01474	4.32%
031.42	1		0.55850														
031.43	3	3	0.57417	0.03155	0.57417	0.03155	0.01288	5.50%	0.03150	0.00261	0.03161	5.49%	0.46%	5.50%	12.091	0.00300	4.35%
031.44	1		0.59550														
032.31	3	3	0.36942	0.01707	0.36942	0.01707	0.00697	4.62%	0.01557	0.00990	0.01845	4.22%	2.68%	5.00%	1.8635	0.01370	4.65%
032.41	8	8	0.35886	0.01506	0.35937	0.01609	0.00402	4.48%	0.00828	0.01779	0.01962	2.31%	4.96%	5.47%	1.1030	0.01780	4.67%
032.42	1		0.39500														
032.43	3	3	0.39933	0.00983	0.39933	0.00983	0.00401	2.46%	0.00931	0.00443	0.01032	2.33%	1.11%	2.58%	2.3262	0.00533	4.59%
032.44	1		0.36550														
033.01	3	3	1.3639	0.07624	1.3639	0.07624	0.03112	5.59%	0.07573	0.01241	0.07674	5.55%	0.91%	5.63%	6.1852	0.01290	3.82%
033.99	1		1.6400														
034.04	1		0.88100														
034.43	1		2.2000														
035.31	3	3	0.60412	0.05594	0.60412	0.05594	0.02284	9.26%	0.05457	0.01736	0.05727	9.03%	2.87%	9.48%	3.2986	0.02097	4.31%
035.41	8	7	0.62885	0.02823	0.62440	0.02145	0.00573	3.44%	0.02789	0.00620	0.02857	4.43%	0.99%	4.54%	4.6108	0.00704	4.29%
035.42	1		0.60850														
035.43	3	2	0.61450	0.01768	0.63750	0.04175	0.01705	6.55%	0.01739	0.00447	0.01796	2.83%	0.73%	2.92%	4.0156	0.02633	4.28%
036.42	1		0.94000														
036.43	1		0.54000														
037.31	2	0															
037.32	1		36.490														
037.33	1		29.595														
037.41	9	8	35.960	3.7458	36.927	5.1593	1.2161	13.97%	3.2499	2.6339	4.1832	9.04%	7.32%	11.63%	1.5882	3.0163	9.29%
037.43	3	3	42.050	10.213	42.050	10.213	4.1696	24.29%	8.9860	6.8650	11.308	21.37%	16.33%	26.89%	1.6472	7.1000	9.11%
037.44	1		36.300														
038.41	1		0.42600														
038.43	1		0.78000														
101.01	1		4,235.0														
103.01	1		8.6750														
104.00	1		7.7850														
105.01	1		0.66000														
106.00	1		1.6000														
106.02	3	2	0.83750	0.09546	390.56	675.02	275.57	172.83%		0.21006			25.08%			33.477	
107.00	1		28.395														
108.02	1		1.2250														
108.99	1		1,230.0														
109.02	1		46.550														

Method Code	Total # Labs Submitting	# Labs Included in Calculations	Mean	SD	Assigned Value - Robust Mean	AAFCCO CS ffp - Robust sd	Uncertainty (U) - Robust	% RSD - Robust	Between Labs sL	Within Labs sr	Reproducibility sR	Between Labs %RSD	Within Labs %/rsd	Reproducibility %RSD	sR/sr	Average Range (R-bar)	Horwitz %RSD
112.00	1		2.7000														
113.01	1		1.5750														
114.01	1		0.99500														
120.00	2	0															
120.05	1		3.0975														
121.00	2	0															
121.05	1		3.4050														
122.00	2	0															
122.05	1		5.7015														
124.00	2	0															
124.05	1		0.79700														
125.00	2	0															
125.05	1		7.3250														
126.00	2	0															
126.05	1		1.9520														
127.00	2	0															
127.05	1		1.2835														
128.00	2	0															
128.05	1		2.5710														
129.00	2	0															
129.05	1		4.5570														
130.00	2	0															
130.05	1		3.7945														
131.00	2	0															
131.05	1		1.9545														
132.00	2	0															
132.05	1		3.1705														
133.00	2	0															
133.05	1		2.0715														
134.00	2	0															
134.05	1		4.2605														
135.00	2	0															
135.05	1		2.6155														
136.00	1		0.99000														
136.05	1		0.65950														
136.99	1		0.95500														
137.00	2	0															
137.05	1		2.2725														

Method Code	Total # Labs Submitting	# Labs Included in Calculations	Mean	SD	Assigned Value - Robust Mean	AAFCCO CS ffp - Robust sd	Uncertainty (U) - Robust	% RSD - Robust	Between Labs sL	Within Labs sr	Reproducibility sR	Between Labs %RSD	Within Labs %/rsd	Reproducibility %RSD	sR/sr	Average Range (R-bar)	Horwitz %RSD
138.00	2	0															
138.05	1		3.3875														
139.00	1		0.03000														
400.01	1		0.28000														
516.00	1		0.01600														
516.53	1		0.03000														
518.41	1		0.00750														
518.43	1		0.00000														
520.41	1		0.54500														
520.43	1		1.5200														
526.41	1		0.01300														
526.43	1		0.38000														
526.53	1		0.10000														
529.99	2	0															
539.41	1		0.00600														
539.43	1		1.3800														
640.01	1		77.500														