



Mycotoxin Contaminants Scheme

Game Bird Feed

Test Material Code # 202064

Method Proficiency Testing Report

Labs Reporting: 53

Methods Reported 81

Issue Date : 11/30/2020

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	2113	6.750	0.9000	10.16	3.513	1.085	12	-0.97	17%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0554	7.450	0.7000	10.16	3.513	1.085	12	-0.77	13%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0027	8.200	0.8000	10.16	3.513	1.085	12	-0.56	10%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0529	9.150	0.3000	10.16	3.513	1.085	12	-0.29	5%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0629	9.185	0.3700	10.16	3.513	1.085	12	-0.28	5%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0511	9.200	0.0000	10.16	3.513	1.085	12	-0.27	5%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0297	9.850	0.7000	10.16	3.513	1.085	12	-0.09	2%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0034	10.98	0.3500	10.16	3.513	1.085	12	0.23	4%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	2096	12.20	6.000	10.16	3.513	1.085	12	0.58	10%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	2098	12.85	0.5000	10.16	3.513	1.085	12	0.77	13%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0876	12.95	0.9000	10.16	3.513	1.085	12	0.79	14%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0208	13.15	1.500	10.16	3.513	1.085	12	0.85	15%	0
600.04	Total Aflatoxin, Charm ROSA Fast 5 Aflatoxin Quant (ppb)	2147	5.725	0.5500				1			0
600.12	Total Aflatoxin, r-Biopharm Ridascreen FAST Afl (ppb)	0202	5.340	0.0000	9.190	3.187	0.3210	4			0
600.12	Total Aflatoxin, r-Biopharm Ridascreen FAST Afl (ppb)	2226	6.889	0.4840	9.190	3.187	0.3210	4			0
600.12	Total Aflatoxin, r-Biopharm Ridascreen FAST Afl (ppb)	0590	11.90	0.8000	9.190	3.187	0.3210	4			0
600.12	Total Aflatoxin, r-Biopharm Ridascreen FAST Afl (ppb)	0656	12.63	0.0000	9.190	3.187	0.3210	4			0
600.13	Total Aflatoxin, r-Biopharm Ridascreen FAST Afl SC (ppb)	0035	8.145	0.5700				3			0
600.13	Total Aflatoxin, r-Biopharm Ridascreen FAST Afl SC (ppb)	0066	9.235	0.0700				3			0
600.13	Total Aflatoxin, r-Biopharm Ridascreen FAST Afl SC (ppb)	0407	9.870	1.460				3			0
600.14	Total Aflatoxin, Vicam Aflatest (ppb)	0190	4.450	0.1000				3			0
600.14	Total Aflatoxin, Vicam Aflatest (ppb)	0004	9.650	0.3000				3			0
600.14	Total Aflatoxin, Vicam Aflatest (ppb)	2268	14.00	0.0000				3			0
600.20	Total Aflatoxin, LC (ppb)	2192	8.645	2.010	12.55	4.316	1.647	4			0
600.20	Total Aflatoxin, LC (ppb)	2244	11.79	0.0370	12.55	4.316	1.647	4			0
600.20	Total Aflatoxin, LC (ppb)	0629	14.02	0.2400	12.55	4.316	1.647	4			0
600.20	Total Aflatoxin, LC (ppb)	0042	15.75	4.300	12.55	4.316	1.647	4			0
600.20	Total Aflatoxin, LC (ppb)	2308	35.49	0.2000	12.55	4.316	1.647	4			2
600.22	Total Aflatoxin, Vicam Aflatest, LC-FI (ppb)	0098	10.80	2.300				1			0
600.24	Total Aflatoxin, LC-MS/MS (ppb)	0682	6.700	0.0000	13.53	4.642	2.132	7	-1.47	25%	0
600.24	Total Aflatoxin, LC-MS/MS (ppb)	0619	6.995	0.0300	13.53	4.642	2.132	7	-1.41	24%	0
600.24	Total Aflatoxin, LC-MS/MS (ppb)	0160	12.25	3.900	13.53	4.642	2.132	7	-0.28	5%	0
600.24	Total Aflatoxin, LC-MS/MS (ppb)	2033	14.60	7.400	13.53	4.642	2.132	7	0.23	4%	0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
600.24	Total Aflatoxin, LC-MS/MS (ppb)	0353	15.46	1.243	13.53	4.642	2.132	7	0.42	7%	0
600.24	Total Aflatoxin, LC-MS/MS (ppb)	2129	18.00	0.8000	13.53	4.642	2.132	7	0.96	17%	0
600.24	Total Aflatoxin, LC-MS/MS (ppb)	0218	20.69	1.552	13.53	4.642	2.132	7	1.54	26%	0
600.98	Total Aflatoxin, Other Rapid Test Kit (ppb)	2269	8.700	0.0000	12.79	4.396	0.1025	4			0
600.98	Total Aflatoxin, Other Rapid Test Kit (ppb)	2244	11.62	0.1100	12.79	4.396	0.1025	4			0
600.98	Total Aflatoxin, Other Rapid Test Kit (ppb)	0142	15.35	0.1000	12.79	4.396	0.1025	4			0
600.98	Total Aflatoxin, Other Rapid Test Kit (ppb)	0870	15.50	0.2000	12.79	4.396	0.1025	4			0
600.98	Total Aflatoxin, Other Rapid Test Kit (ppb)	2169	9.900	0.8000	12.79	4.396	0.1025	4			1
601.20	Aflatoxin B1, LC (ppb)	2192	7.685	1.730	14.61	5.003	0.8953	4			0
601.20	Aflatoxin B1, LC (ppb)	2244	10.80	0.1510	14.61	5.003	0.8953	4			0
601.20	Aflatoxin B1, LC (ppb)	0629	12.85	0.3000	14.61	5.003	0.8953	4			0
601.20	Aflatoxin B1, LC (ppb)	2308	27.10	1.400	14.61	5.003	0.8953	4			0
601.21	Aflatoxin B1, LC-PCD FI (ppb)	0868	7.030	0.1400				2			0
601.21	Aflatoxin B1, LC-PCD FI (ppb)	0910	16.91	0.7200				2			0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0682	6.200	0.0000	12.53	4.307	2.556	14	-1.47	25%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0619	6.995	0.0300	12.53	4.307	2.556	14	-1.28	22%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	2055	8.850	0.7000	12.53	4.307	2.556	14	-0.85	15%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0870	9.090	1.420	12.53	4.307	2.556	14	-0.80	14%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	2136	9.400	3.200	12.53	4.307	2.556	14	-0.73	12%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0160	11.70	2.800	12.53	4.307	2.556	14	-0.19	3%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0043	12.04	6.312	12.53	4.307	2.556	14	-0.11	2%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	2033	13.40	6.600	12.53	4.307	2.556	14	0.20	3%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0553	13.60	0.8100	12.53	4.307	2.556	14	0.25	4%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0353	14.47	1.027	12.53	4.307	2.556	14	0.45	8%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	2327	15.36	3.154	12.53	4.307	2.556	14	0.66	11%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0964	16.38	3.256	12.53	4.307	2.556	14	0.89	15%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	2140	18.50	5.000	12.53	4.307	2.556	14	1.39	24%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0218	19.39	1.474	12.53	4.307	2.556	14	1.59	27%	0
601.24	Aflatoxin B1, Vicam Aflatest, LC-FI (ppb)	0098	10.21	1.120				2			0
601.24	Aflatoxin B1, Vicam Aflatest, LC-FI (ppb)	2053	10.46	2.040				2			0
601.98	Aflatoxin B1, Other Rapid Test Kit (ppb)	0868	6.600	0.6000	8.999	3.122	0.4075	4			0
601.98	Aflatoxin B1, Other Rapid Test Kit (ppb)	2239	7.550	0.3000	8.999	3.122	0.4075	4			0
601.98	Aflatoxin B1, Other Rapid Test Kit (ppb)	2244	10.75	0.1300	8.999	3.122	0.4075	4			0
601.98	Aflatoxin B1, Other Rapid Test Kit (ppb)	2269	11.10	0.6000	8.999	3.122	0.4075	4			0
602.20	Aflatoxin B2, LC (ppb)	2192	0.5750	0.1100				3			0
602.20	Aflatoxin B2, LC (ppb)	2244	0.9790	0.1860				3			0
602.20	Aflatoxin B2, LC (ppb)	0629	1.170	0.0600				3			0
602.20	Aflatoxin B2, LC (ppb)	2308	3.090	1.520				3			1
602.21	Aflatoxin B2, LC-PCD FI (ppb)	0910	1.260	0.1600				1			0
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0682	0.5000	0.0000	1.152	0.4227	0.2329	9	-1.54	28%	0
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0353	0.7215	0.0530	1.152	0.4227	0.2329	9	-1.02	19%	0
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0870	0.7300	0.1000	1.152	0.4227	0.2329	9	-1.00	18%	0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
602.23	Aflatoxin B2, LC-MS/MS (ppb)	2136	0.9100	0.0600	1.152	0.4227	0.2329	9	-0.57	11%	0
602.23	Aflatoxin B2, LC-MS/MS (ppb)	2327	1.105	0.2110	1.152	0.4227	0.2329	9	-0.11	2%	0
602.23	Aflatoxin B2, LC-MS/MS (ppb)	2033	1.200	0.8000	1.152	0.4227	0.2329	9	0.11	2%	0
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0218	1.303	0.0783	1.152	0.4227	0.2329	9	0.36	7%	0
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0964	1.865	0.4240	1.152	0.4227	0.2329	9	1.69	31%	0
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0553	2.135	0.3700	1.152	0.4227	0.2329	9	2.32	43%	0
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0619	0.0000	0.0000	1.152	0.4227	0.2329	9			4
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0043	< 0.6		1.152	0.4227	0.2329	9			5
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0160	< 1		1.152	0.4227	0.2329	9			5
602.23	Aflatoxin B2, LC-MS/MS (ppb)	2140	< 20		1.152	0.4227	0.2329	9			5
602.24	Aflatoxin B2, Vicam Aflatest, LC-FI (ppb)	2053	0.7850	0.0100				1			0
602.24	Aflatoxin B2, Vicam Aflatest, LC-FI (ppb)	0098	< 1					1			5
603.20	Aflatoxin G1, LC (ppb)	2308	0.3750	0.7500				0			4
603.20	Aflatoxin G1, LC (ppb)	0629	< 0.25					0			5
603.21	Aflatoxin G1, LC-PCD FI (ppb)	0910	0.1250	0.1700				1			0
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0619	0.0000	0.0000				0			4
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0682	< 0.5					0			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0870	< 0.6					0			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	2136	< 0.83					0			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0160	< 1					0			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0218	< 1					0			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0553	< 1					0			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0964	< 1					0			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	2327	< 1.1					0			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0043	< 4					0			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	2033	< 5					0			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	2140	< 20					0			5
603.24	Aflatoxin G1, Vicam Aflatest, LC-FI (ppb)	2053	< 0.32					0			5
603.24	Aflatoxin G1, Vicam Aflatest, LC-FI (ppb)	0098	< 15					0			5
604.20	Aflatoxin G2, LC (ppb)	2308	4.925	1.070				1			0
604.20	Aflatoxin G2, LC (ppb)	0629	< 0.25					1			5
604.21	Aflatoxin G2, LC-PCD FI (ppb)	0910	< 0.01					0			5
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0619	0.0000	0.0000				0			4
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0682	< 0.5					0			5
604.23	Aflatoxin G2, LC-MS/MS (ppb)	2136	< 0.83					0			5
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0160	< 1					0			5
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0218	< 1					0			5
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0553	< 1					0			5
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0870	< 1					0			5
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0964	< 1					0			5
604.23	Aflatoxin G2, LC-MS/MS (ppb)	2327	< 1.6					0			5
604.23	Aflatoxin G2, LC-MS/MS (ppb)	2033	< 5					0			5

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
604.23	Aflatoxin G2, LC-MS/MS (ppb)	2140	< 20					0		5	
604.24	Aflatoxin G2, Vicam Aflatest, LC-Fl (ppb)	2053	< 0.49					0		5	
604.24	Aflatoxin G2, Vicam Aflatest, LC-Fl (ppb)	0098	< 5					0		5	
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0027	1,200	200.0	1,432	432.9	122.5	12	-0.54	8%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	2192	1,280	320.0	1,432	432.9	122.5	12	-0.35	5%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0297	1,350	100.0	1,432	432.9	122.5	12	-0.19	3%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0529	1,350	300.0	1,432	432.9	122.5	12	-0.19	3%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	2098	1,360	80.00	1,432	432.9	122.5	12	-0.17	3%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0208	1,395	70.00	1,432	432.9	122.5	12	-0.08	1%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0511	1,400	0.0000	1,432	432.9	122.5	12	-0.07	1%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0870	1,400	0.0000	1,432	432.9	122.5	12	-0.07	1%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0042	1,450	100.0	1,432	432.9	122.5	12	0.04	1%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0554	1,650	100.0	1,432	432.9	122.5	12	0.50	8%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	2113	1,650	100.0	1,432	432.9	122.5	12	0.50	8%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0413	1,750	100.0	1,432	432.9	122.5	12	0.74	11%	0
610.04	Deoxynivalenol, Charm ROSA Fast 5 DON Quant (ppb)	2147	1,225	150.0				1		0	
610.06	Deoxynivalenol, r-Biopharm Ridascreen DON (ppb)	0202	1,140	45.00				1		0	
610.07	Deoxynivalenol, r-Biopharm Ridascreen FAST DON (ppb)	0638	987.5	55.00	1,299	393.8	76.20	5		0	
610.07	Deoxynivalenol, r-Biopharm Ridascreen FAST DON (ppb)	0590	1,255	110.0	1,299	393.8	76.20	5		0	
610.07	Deoxynivalenol, r-Biopharm Ridascreen FAST DON (ppb)	0407	1,344	16.00	1,299	393.8	76.20	5		0	
610.07	Deoxynivalenol, r-Biopharm Ridascreen FAST DON (ppb)	0629	1,440	200.0	1,299	393.8	76.20	5		0	
610.07	Deoxynivalenol, r-Biopharm Ridascreen FAST DON (ppb)	0656	1,467	0.0000	1,299	393.8	76.20	5		0	
610.08	Deoxynivalenol, r-Biopharm Ridascreen FAST DON SC (p	0066	1,355	250.0				2		0	
610.08	Deoxynivalenol, r-Biopharm Ridascreen FAST DON SC (p	0035	2,044	62.00				2		0	
610.20	Deoxynivalenol, LC (ppb)	0910	1,240	0.0600				3		0	
610.20	Deoxynivalenol, LC (ppb)	2244	1,433	10.13				3		0	
610.20	Deoxynivalenol, LC (ppb)	0413	1,600	0.0000				3		0	
610.20	Deoxynivalenol, LC (ppb)	0098	1,150	100.0				3		1	
610.22	Deoxynivalenol, LC-MS (ppb)	2268	1,314	160.0				1		0	
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0218	1,180	0.2000	1,087	331.3	66.16	16	-3.28	50%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0160	1,450	0.1000	1,087	331.3	66.16	16	-3.28	50%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0043	533.1	41.90	1,087	331.3	66.16	16	-1.67	25%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	2053	762.0	0.9200	1,087	331.3	66.16	16	-0.98	15%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	2136	955.0	70.00	1,087	331.3	66.16	16	-0.40	6%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	2033	973.5	29.00	1,087	331.3	66.16	16	-0.34	5%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0353	996.7	55.44	1,087	331.3	66.16	16	-0.27	4%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0619	1,053	71.00	1,087	331.3	66.16	16	-0.11	2%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	2140	1,155	11.00	1,087	331.3	66.16	16	0.20	3%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0682	1,210	0.0000	1,087	331.3	66.16	16	0.37	6%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	2129	1,216	22.00	1,087	331.3	66.16	16	0.39	6%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0870	1,240	80.00	1,087	331.3	66.16	16	0.46	7%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0553	1,381	198.5	1,087	331.3	66.16	16	0.89	14%	0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
610.23	Deoxynivalenol, LC-MS/MS (ppb)	2327	1,598	83.16	1,087	331.3	66.16	16	1.54	23%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	2055	1,895	380.9	1,087	331.3	66.16	16	2.44	37%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0964	2,014	14.40	1,087	331.3	66.16	16	2.80	43%	0
610.98	Deoxynivalenol, Other Rapid Test Kit (ppb)	0868	1,120	20.00	1,362	412.5	53.55	6	-0.59	9%	0
610.98	Deoxynivalenol, Other Rapid Test Kit (ppb)	2269	1,135	78.00	1,362	412.5	53.55	6	-0.55	8%	0
610.98	Deoxynivalenol, Other Rapid Test Kit (ppb)	2239	1,195	12.00	1,362	412.5	53.55	6	-0.41	6%	0
610.98	Deoxynivalenol, Other Rapid Test Kit (ppb)	2244	1,423	11.27	1,362	412.5	53.55	6	0.15	2%	0
610.98	Deoxynivalenol, Other Rapid Test Kit (ppb)	0142	1,650	100.0	1,362	412.5	53.55	6	0.70	11%	0
610.98	Deoxynivalenol, Other Rapid Test Kit (ppb)	2169	1,650	100.0	1,362	412.5	53.55	6	0.70	11%	0
610.99	Deoxynivalenol, Miscellaneous (ppb)	0870	920.0	120.0				2			0
610.99	Deoxynivalenol, Miscellaneous (ppb)	0037	1,300	0.0000				2			0
620.01	Total Fumonisin, Neogen Veratox Fumonisin (ppb)	0554	2,550	100.0	3,763	1,109	441.7	6	-1.09	16%	0
620.01	Total Fumonisin, Neogen Veratox Fumonisin (ppb)	0042	3,100	400.0	3,763	1,109	441.7	6	-0.60	9%	0
620.01	Total Fumonisin, Neogen Veratox Fumonisin (ppb)	0511	3,450	100.0	3,763	1,109	441.7	6	-0.28	4%	0
620.01	Total Fumonisin, Neogen Veratox Fumonisin (ppb)	2113	3,950	100.0	3,763	1,109	441.7	6	0.17	2%	0
620.01	Total Fumonisin, Neogen Veratox Fumonisin (ppb)	0027	4,055	1,150	3,763	1,109	441.7	6	0.26	4%	0
620.01	Total Fumonisin, Neogen Veratox Fumonisin (ppb)	2192	6,450	800.0	3,763	1,109	441.7	6	2.42	36%	0
620.02	Total Fumonisin, Neogen Veratox Fumonisin 5/10 (ppb)	2098	2,565	190.0				3			0
620.02	Total Fumonisin, Neogen Veratox Fumonisin 5/10 (ppb)	0870	3,595	130.0				3			0
620.02	Total Fumonisin, Neogen Veratox Fumonisin 5/10 (ppb)	0529	3,950	100.0				3			0
620.04	Total Fumonisin, Charm ROSA Fast 5 FUMQ (ppb)	2147	5,125	1,450				1			0
620.07	Total Fumonisin, Romer AgraQuant Total Fumonisin (ppb)	0868	3,180	0.0000				1			0
620.09	Total Fumonisin, r-Biopharm Ridascreen FAST FUM (ppb)	0656	2,508	0.0000	3,537	1,044	211.5	6	-0.99	15%	0
620.09	Total Fumonisin, r-Biopharm Ridascreen FAST FUM (ppb)	2226	2,711	18.00	3,537	1,044	211.5	6	-0.79	12%	0
620.09	Total Fumonisin, r-Biopharm Ridascreen FAST FUM (ppb)	0407	2,873	405.0	3,537	1,044	211.5	6	-0.64	9%	0
620.09	Total Fumonisin, r-Biopharm Ridascreen FAST FUM (ppb)	0202	3,430	148.0	3,537	1,044	211.5	6	-0.10	2%	0
620.09	Total Fumonisin, r-Biopharm Ridascreen FAST FUM (ppb)	0066	4,610	660.0	3,537	1,044	211.5	6	1.03	15%	0
620.09	Total Fumonisin, r-Biopharm Ridascreen FAST FUM (ppb)	0035	5,092	38.00	3,537	1,044	211.5	6	1.49	22%	0
620.11	Total Fumonisin, Vicam Fumonitest (ppb)	0098	1,150	100.0				3			0
620.11	Total Fumonisin, Vicam Fumonitest (ppb)	2268	2,500	400.0				3			0
620.11	Total Fumonisin, Vicam Fumonitest (ppb)	0190	2,550	300.0				3			0
620.20	Total Fumonisin, LC (ppb)	0208	4,150	500.0				1			0
620.23	Total Fumonisin, LC-MS/MS (ppb)	0160	3,050	0.3000	3,032	898.4	256.1	6	-3.37	50%	0
620.23	Total Fumonisin, LC-MS/MS (ppb)	0619	2,773	375.0	3,032	898.4	256.1	6	-0.29	4%	0
620.23	Total Fumonisin, LC-MS/MS (ppb)	0353	3,114	615.0	3,032	898.4	256.1	6	0.09	1%	0
620.23	Total Fumonisin, LC-MS/MS (ppb)	0682	3,230	0.0000	3,032	898.4	256.1	6	0.22	3%	0
620.23	Total Fumonisin, LC-MS/MS (ppb)	2244	3,319	351.5	3,032	898.4	256.1	6	0.32	5%	0
620.23	Total Fumonisin, LC-MS/MS (ppb)	2033	4,310	195.0	3,032	898.4	256.1	6	1.42	21%	0
620.98	Total Fumonisin, Other Rapid Test Kit (ppb)	2239	3,063	26.00				3			0
620.98	Total Fumonisin, Other Rapid Test Kit (ppb)	0868	3,130	20.00				3			0
620.98	Total Fumonisin, Other Rapid Test Kit (ppb)	0142	4,900	0.0000				3			0
620.98	Total Fumonisin, Other Rapid Test Kit (ppb)	2269	4,150	900.0				3			1

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
621.20	Fumonsin B1, LC (ppb)	2244	2,648	237.5				1			0
621.21	Fumonsin B1, LC-FI OPA der. (ppb)	0098	950.0	100.0				2			0
621.21	Fumonsin B1, LC-FI OPA der. (ppb)	0910	2,154	208.0				2			0
621.22	Fumonsin B1, LC-MS (ppb)	2053	1,351	334.2				1			0
621.23	Fumonsin B1, LC-MS/MS (ppb)	0218	2.205	0.1500	1,875	562.8	84.79	14	-3.33	50%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	0160	2.450	0.3000	1,875	562.8	84.79	14	-3.33	50%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	2055	922.6	240.0	1,875	562.8	84.79	14	-1.69	25%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	0043	1,152	51.00	1,875	562.8	84.79	14	-1.28	19%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	0964	1,754	34.91	1,875	562.8	84.79	14	-0.21	3%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	0553	1,842	55.10	1,875	562.8	84.79	14	-0.06	1%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	0619	2,186	224.0	1,875	562.8	84.79	14	0.55	8%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	0353	2,218	105.3	1,875	562.8	84.79	14	0.61	9%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	0870	2,320	0.0000	1,875	562.8	84.79	14	0.79	12%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	2136	2,400	200.0	1,875	562.8	84.79	14	0.93	14%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	2140	2,419	163.0	1,875	562.8	84.79	14	0.97	15%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	2327	2,458	33.29	1,875	562.8	84.79	14	1.04	16%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	0682	2,530	0.0000	1,875	562.8	84.79	14	1.16	17%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	2033	2,920	80.00	1,875	562.8	84.79	14	1.86	28%	0
622.20	Fumonisin B2, LC (ppb)	2244	670.4	114.0				1			0
622.21	Fumonisin B2, LC-FI OPA der. (ppb)	0098	200.0	0.0000				2			0
622.21	Fumonisin B2, LC-FI OPA der. (ppb)	0910	550.5	177.0				2			0
622.22	Fumonisin B2, LC-MS (ppb)	2053	397.6	59.30				1			0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0160	0.5000	0.0000	573.1	177.7	100.2	14	-3.22	50%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0218	0.5125	0.0430	573.1	177.7	100.2	14	-3.22	50%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0043	314.6	32.00	573.1	177.7	100.2	14	-1.45	23%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	2136	500.0	40.00	573.1	177.7	100.2	14	-0.41	6%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	2140	526.0	20.00	573.1	177.7	100.2	14	-0.26	4%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	2327	527.6	20.35	573.1	177.7	100.2	14	-0.26	4%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0619	586.5	151.0	573.1	177.7	100.2	14	0.08	1%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0870	620.0	0.0000	573.1	177.7	100.2	14	0.26	4%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0964	626.7	12.45	573.1	177.7	100.2	14	0.30	5%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0353	659.5	37.59	573.1	177.7	100.2	14	0.49	8%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0553	697.0	75.20	573.1	177.7	100.2	14	0.70	11%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0682	700.0	0.0000	573.1	177.7	100.2	14	0.71	11%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	2033	989.5	101.0	573.1	177.7	100.2	14	2.34	36%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	2055	3,356	913.6	573.1	177.7	100.2	14	15.67	243%	0
623.22	Fumonisin B3, LC-MS (ppb)	2053	150.4	30.22				1			0
623.23	Fumonisin B3, LC-MS/MS (ppb)	0160	0.1000	0.0000	257.8	81.68	9.872	5			0
623.23	Fumonisin B3, LC-MS/MS (ppb)	2327	244.4	6.662	257.8	81.68	9.872	5			0
623.23	Fumonisin B3, LC-MS/MS (ppb)	0870	280.0	0.0000	257.8	81.68	9.872	5			0
623.23	Fumonisin B3, LC-MS/MS (ppb)	0553	364.7	28.70	257.8	81.68	9.872	5			0
623.23	Fumonisin B3, LC-MS/MS (ppb)	2033	400.0	14.00	257.8	81.68	9.872	5			0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
630.01	Ochratoxin A, Neogen Veratox Ochratoxin (ppb)	2098	16.08	0.4300				2			0
630.01	Ochratoxin A, Neogen Veratox Ochratoxin (ppb)	0529	117.8	0.5000				2			0
630.03	Ochratoxin A, Romer AgraQuant Ochratoxin (ppb)	0868	18.36	0.5900				1			0
630.04	Ochratoxin A, r-Biopharm Ridascreen Ochratoxin A (ppb)	0629	9.565	0.0100				2			0
630.04	Ochratoxin A, r-Biopharm Ridascreen Ochratoxin A (ppb)	0407	19.23	1.750				2			0
630.05	Ochratoxin A, r-Biopharm Ridascreen FAST Och A (ppb)	2226	7.161	0.6250				3			0
630.05	Ochratoxin A, r-Biopharm Ridascreen FAST Och A (ppb)	0066	21.67	0.3300				3			0
630.05	Ochratoxin A, r-Biopharm Ridascreen FAST Och A (ppb)	0656	24.45	0.0000				3			0
630.20	Ochratoxin A, LC (ppb)	2244	16.45	0.2340				3			0
630.20	Ochratoxin A, LC (ppb)	0098	27.25	1.700				3			0
630.20	Ochratoxin A, LC (ppb)	2192	32.97	11.41				3			0
630.21	Ochratoxin A, LC-PCD FI (ppb)	0910	19.75	2.300				1			0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0870	11.12	0.2800	17.26	5.882	1.878	15	-1.04	18%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0043	11.65	0.6300	17.26	5.882	1.878	15	-0.95	16%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0682	12.40	0.0000	17.26	5.882	1.878	15	-0.83	14%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0619	12.80	1.000	17.26	5.882	1.878	15	-0.76	13%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0553	13.10	3.770	17.26	5.882	1.878	15	-0.71	12%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	2053	13.70	0.2300	17.26	5.882	1.878	15	-0.61	10%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0353	16.47	0.3570	17.26	5.882	1.878	15	-0.13	2%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	2327	17.09	1.475	17.26	5.882	1.878	15	-0.03	0%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	2136	19.50	5.000	17.26	5.882	1.878	15	0.38	7%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	2033	20.50	5.000	17.26	5.882	1.878	15	0.55	9%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	2055	20.70	0.6000	17.26	5.882	1.878	15	0.59	10%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0160	21.50	5.000	17.26	5.882	1.878	15	0.72	12%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	2140	21.50	3.000	17.26	5.882	1.878	15	0.72	12%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0964	22.73	1.100	17.26	5.882	1.878	15	0.93	16%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0218	24.09	0.7354	17.26	5.882	1.878	15	1.16	20%	0
630.98	Ochratoxin A, Other Rapid Test Kit (ppb)	0042	9.928	4.188	19.42	6.600	1.554	6	-1.44	24%	0
630.98	Ochratoxin A, Other Rapid Test Kit (ppb)	2244	16.50	0.0230	19.42	6.600	1.554	6	-0.44	8%	0
630.98	Ochratoxin A, Other Rapid Test Kit (ppb)	2239	16.65	0.3000	19.42	6.600	1.554	6	-0.42	7%	0
630.98	Ochratoxin A, Other Rapid Test Kit (ppb)	0868	18.40	1.810	19.42	6.600	1.554	6	-0.16	3%	0
630.98	Ochratoxin A, Other Rapid Test Kit (ppb)	0870	22.85	1.300	19.42	6.600	1.554	6	0.52	9%	0
630.98	Ochratoxin A, Other Rapid Test Kit (ppb)	2269	48.55	1.700	19.42	6.600	1.554	6	4.41	75%	0
640.03	T-2 Toxin, Romer AgraQuant T-2 (ppb)	0066	97.45	2.300				2			0
640.03	T-2 Toxin, Romer AgraQuant T-2 (ppb)	0868	130.7	0.5700				2			0
640.05	T-2 Toxin, r-Biopharm Ridascreen FAST T-2 (ppb)	0638	72.50	5.000				3			0
640.05	T-2 Toxin, r-Biopharm Ridascreen FAST T-2 (ppb)	0407	128.4	17.94				3			0
640.05	T-2 Toxin, r-Biopharm Ridascreen FAST T-2 (ppb)	0656	138.5	0.0000				3			0
640.21	T-2 Toxin, LC-PCD FI (ppb)	0910	55.50	1.000				1			0
640.23	T-2 Toxin, LC-MS/MS (ppb)	0218	0.0885	0.0330	70.94	23.27	6.970	14	-3.04	50%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	0553	23.24	13.32	70.94	23.27	6.970	14	-2.05	34%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	2053	55.91	1.930	70.94	23.27	6.970	14	-0.65	11%	0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
640.23	T-2 Toxin, LC-MS/MS (ppb)	0619	63.45	0.3000	70.94	23.27	6.970	14	-0.32	5%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	2055	67.40	17.80	70.94	23.27	6.970	14	-0.15	2%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	0964	72.04	5.017	70.94	23.27	6.970	14	0.05	1%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	0353	72.09	1.678	70.94	23.27	6.970	14	0.05	1%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	0682	72.20	0.0000	70.94	23.27	6.970	14	0.05	1%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	2129	74.00	4.000	70.94	23.27	6.970	14	0.13	2%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	2136	74.00	8.000	70.94	23.27	6.970	14	0.13	2%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	2140	82.50	1.000	70.94	23.27	6.970	14	0.50	8%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	2033	85.00	30.00	70.94	23.27	6.970	14	0.60	10%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	0870	85.96	8.920	70.94	23.27	6.970	14	0.65	11%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	2244	104.5	5.577	70.94	23.27	6.970	14	1.44	24%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	0160	< 0.1		70.94	23.27	6.970	14			5
640.23	T-2 Toxin, LC-MS/MS (ppb)	2327	< 200		70.94	23.27	6.970	14			5
640.98	T-2 Toxin, Other Rapid Test Kit (ppb)	2239	107.5	3.000				3			0
640.98	T-2 Toxin, Other Rapid Test Kit (ppb)	0868	112.7	2.420				3			0
640.98	T-2 Toxin, Other Rapid Test Kit (ppb)	2269	197.1	5.600				3			0
640.98	T-2 Toxin, Other Rapid Test Kit (ppb)	0042	135.2	73.70				3			1
641.21	HT-2 Toxin, LC post-col photochem der. FI (ppb)	0910	71.50	3.000				1			0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	0218	0.1175	0.0050	107.7	34.95	10.00	10	-3.08	50%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	0619	90.55	4.100	107.7	34.95	10.00	10	-0.49	8%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	2033	95.50	15.00	107.7	34.95	10.00	10	-0.35	6%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	2053	98.70	11.83	107.7	34.95	10.00	10	-0.26	4%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	0870	106.8	11.14	107.7	34.95	10.00	10	-0.03	0%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	0353	109.7	12.49	107.7	34.95	10.00	10	0.06	1%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	2136	110.0	0.0000	107.7	34.95	10.00	10	0.06	1%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	0964	127.5	6.937	107.7	34.95	10.00	10	0.56	9%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	2055	128.9	32.50	107.7	34.95	10.00	10	0.60	10%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	2140	133.0	6.000	107.7	34.95	10.00	10	0.72	12%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	2327	< 200		107.7	34.95	10.00	10			5
642.21	T-2 Toxin + HT-2 Toxin, Neogen Veratox T-2 + HT-2 (ppb)	2192	162.4	25.20	302.9	95.54	23.94	4			0
642.21	T-2 Toxin + HT-2 Toxin, Neogen Veratox T-2 + HT-2 (ppb)	2098	296.5	18.74	302.9	95.54	23.94	4			0
642.21	T-2 Toxin + HT-2 Toxin, Neogen Veratox T-2 + HT-2 (ppb)	0870	350.2	15.60	302.9	95.54	23.94	4			0
642.21	T-2 Toxin + HT-2 Toxin, Neogen Veratox T-2 + HT-2 (ppb)	0529	402.5	36.20	302.9	95.54	23.94	4			0
642.98	T-2 Toxin + HT-2 Toxin, Other Rapid Test Kit (ppb)	0142	152.9	13.40	227.6	72.34	11.34	4			0
642.98	T-2 Toxin + HT-2 Toxin, Other Rapid Test Kit (ppb)	2244	208.9	11.15	227.6	72.34	11.34	4			0
642.98	T-2 Toxin + HT-2 Toxin, Other Rapid Test Kit (ppb)	2269	272.0	7.800	227.6	72.34	11.34	4			0
642.98	T-2 Toxin + HT-2 Toxin, Other Rapid Test Kit (ppb)	0590	276.5	13.00	227.6	72.34	11.34	4			0
642.98	T-2 Toxin + HT-2 Toxin, Other Rapid Test Kit (ppb)	0629	306.5	57.00	227.6	72.34	11.34	4			1
642.99	T-2 Toxin + HT-2 Toxin, Miscellaneous Technique (ppb)	0037	175.8	3.900				2			0
642.99	T-2 Toxin + HT-2 Toxin, Miscellaneous Technique (ppb)	0407	270.3	11.72				2			0
650.01	Zearalenone, Neogen Veratox Zearalenone (ppb)	2098	225.1	36.26	392.6	123.0	53.73	6	-1.36	21%	0
650.01	Zearalenone, Neogen Veratox Zearalenone (ppb)	0870	335.7	11.80	392.6	123.0	53.73	6	-0.46	7%	0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
650.01	Zearalenone, Neogen Veratox Zearalenone (ppb)	2113	362.0	160.0	392.6	123.0	53.73	6	-0.25	4%	0
650.01	Zearalenone, Neogen Veratox Zearalenone (ppb)	0027	377.0	101.2	392.6	123.0	53.73	6	-0.13	2%	0
650.01	Zearalenone, Neogen Veratox Zearalenone (ppb)	0529	443.7	11.50	392.6	123.0	53.73	6	0.42	7%	0
650.01	Zearalenone, Neogen Veratox Zearalenone (ppb)	0554	650.9	1.600	392.6	123.0	53.73	6	2.10	33%	0
650.06	Zearalenone, r-Biopharm Ridascreen FAST ZON (ppb)	0202	181.4	0.8000	254.0	80.51	1.409	4			0
650.06	Zearalenone, r-Biopharm Ridascreen FAST ZON (ppb)	2226	182.6	3.326	254.0	80.51	1.409	4			0
650.06	Zearalenone, r-Biopharm Ridascreen FAST ZON (ppb)	0590	313.5	1.000	254.0	80.51	1.409	4			0
650.06	Zearalenone, r-Biopharm Ridascreen FAST ZON (ppb)	0656	338.6	0.5100	254.0	80.51	1.409	4			0
650.07	Zearalenone, r-Biopharm Ridascreen FAST ZON SC (ppb)	0407	280.7	11.54				3			0
650.07	Zearalenone, r-Biopharm Ridascreen FAST ZON SC (ppb)	0066	306.3	18.10				3			0
650.07	Zearalenone, r-Biopharm Ridascreen FAST ZON SC (ppb)	0035	326.5	11.97				3			0
650.20	Zearalenone, LC (ppb)	0098	170.0	60.00	307.4	96.91	26.52	6	-1.42	22%	0
650.20	Zearalenone, LC (ppb)	0910	205.0	12.00	307.4	96.91	26.52	6	-1.06	17%	0
650.20	Zearalenone, LC (ppb)	0868	263.0	4.000	307.4	96.91	26.52	6	-0.46	7%	0
650.20	Zearalenone, LC (ppb)	0413	310.0	60.00	307.4	96.91	26.52	6	0.03	0%	0
650.20	Zearalenone, LC (ppb)	2244	369.1	22.24	307.4	96.91	26.52	6	0.64	10%	0
650.20	Zearalenone, LC (ppb)	2192	701.0	0.8700	307.4	96.91	26.52	6	4.06	64%	0
650.23	Zearalenone, LC-MS (ppb)	0353	238.2	17.27				2			0
650.23	Zearalenone, LC-MS (ppb)	0160	474.5	309.0				2			0
650.24	Zearalenone, LC-MS/MS (ppb)	2053	138.7	0.0700	242.3	76.88	37.21	14	-1.35	21%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0043	160.7	44.90	242.3	76.88	37.21	14	-1.06	17%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0870	189.9	17.21	242.3	76.88	37.21	14	-0.68	11%	0
650.24	Zearalenone, LC-MS/MS (ppb)	2033	215.0	10.00	242.3	76.88	37.21	14	-0.35	6%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0619	219.5	21.00	242.3	76.88	37.21	14	-0.30	5%	0
650.24	Zearalenone, LC-MS/MS (ppb)	2136	225.0	50.00	242.3	76.88	37.21	14	-0.22	4%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0218	254.5	12.42	242.3	76.88	37.21	14	0.16	3%	0
650.24	Zearalenone, LC-MS/MS (ppb)	2055	255.3	6.900	242.3	76.88	37.21	14	0.17	3%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0964	258.7	10.93	242.3	76.88	37.21	14	0.21	3%	0
650.24	Zearalenone, LC-MS/MS (ppb)	2327	269.4	14.85	242.3	76.88	37.21	14	0.35	6%	0
650.24	Zearalenone, LC-MS/MS (ppb)	2129	286.0	56.00	242.3	76.88	37.21	14	0.57	9%	0
650.24	Zearalenone, LC-MS/MS (ppb)	2140	287.5	9.000	242.3	76.88	37.21	14	0.59	9%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0682	298.0	0.0000	242.3	76.88	37.21	14	0.72	12%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0553	313.6	267.6	242.3	76.88	37.21	14	0.93	15%	0
650.98	Zearalenone, Other Rapid Test Kit (ppb)	2147	205.0	40.00	282.9	89.38	20.82	6	-0.87	14%	0
650.98	Zearalenone, Other Rapid Test Kit (ppb)	2239	212.5	5.000	282.9	89.38	20.82	6	-0.79	12%	0
650.98	Zearalenone, Other Rapid Test Kit (ppb)	0868	230.5	1.000	282.9	89.38	20.82	6	-0.59	9%	0
650.98	Zearalenone, Other Rapid Test Kit (ppb)	2269	335.7	36.30	282.9	89.38	20.82	6	0.59	9%	0
650.98	Zearalenone, Other Rapid Test Kit (ppb)	2244	340.9	20.69	282.9	89.38	20.82	6	0.65	10%	0
650.98	Zearalenone, Other Rapid Test Kit (ppb)	0142	372.7	21.90	282.9	89.38	20.82	6	1.00	16%	0
650.99	Zearalenone, Miscellaneous (ppb)	2268	190.0	20.00				2			0
650.99	Zearalenone, Miscellaneous (ppb)	0037	319.3	62.10				2			0

Notes: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			

and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = Rejected for duplicates too far apart, 2 = Rejected as outlier, 5 = A reporting limit and 4 = zeros submitted as values. Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). Flag 3 indicates not used in statistics.