



Swine Feed, Medicated

Animal Feed Scheme

Labs Reporting: 195

Test Material Code # 201725

Analyte Proficiency Testing Report

Issue Date : 06/30/2017

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
000	Urea (%)	0278	0.30000	0.00000				1		000.02	0
001	Loss on Drying (%)	0618	6.6050	0.09000	9.0213	0.57073	0.13142	74	-4.23	001.99	0
001	Loss on Drying (%)	0618	6.8780	0.00600	9.0213	0.57073	0.13142	74	-3.76	001.07	0
001	Loss on Drying (%)	2155	7.9250	0.05000	9.0213	0.57073	0.13142	74	-1.92	001.99	0
001	Loss on Drying (%)	0034	8.2325	0.07500	9.0213	0.57073	0.13142	74	-1.38	001.00	0
001	Loss on Drying (%)	0297	8.2650	0.03000	9.0213	0.57073	0.13142	74	-1.33	001.07	0
001	Loss on Drying (%)	0536	8.2750	0.03000	9.0213	0.57073	0.13142	74	-1.31	001.99	0
001	Loss on Drying (%)	2149	8.3360	0.02000	9.0213	0.57073	0.13142	74	-1.20	001.99	0
001	Loss on Drying (%)	2012	8.3600	0.06000	9.0213	0.57073	0.13142	74	-1.16	001.07	0
001	Loss on Drying (%)	0630	8.3750	0.03000	9.0213	0.57073	0.13142	74	-1.13	001.99	0
001	Loss on Drying (%)	0675	8.4200	0.32000	9.0213	0.57073	0.13142	74	-1.05	001.07	0
001	Loss on Drying (%)	0407	8.4298	0.13480	9.0213	0.57073	0.13142	74	-1.04	001.07	0
001	Loss on Drying (%)	0610	8.4350	0.05000	9.0213	0.57073	0.13142	74	-1.03	001.05	0
001	Loss on Drying (%)	2093	8.4900	0.04000	9.0213	0.57073	0.13142	74	-0.93	001.03	0
001	Loss on Drying (%)	0049	8.5550	0.17000	9.0213	0.57073	0.13142	74	-0.82	001.07	0
001	Loss on Drying (%)	0278	8.6200	0.34000	9.0213	0.57073	0.13142	74	-0.70	001.07	0
001	Loss on Drying (%)	0889	8.6200	0.30000	9.0213	0.57073	0.13142	74	-0.70	001.99	0
001	Loss on Drying (%)	0918	8.6250	0.09000	9.0213	0.57073	0.13142	74	-0.69	001.99	0
001	Loss on Drying (%)	0895	8.6450	0.01000	9.0213	0.57073	0.13142	74	-0.66	001.03	0
001	Loss on Drying (%)	0510	8.6500	0.10000	9.0213	0.57073	0.13142	74	-0.65	001.99	0
001	Loss on Drying (%)	0948	8.6500	0.10000	9.0213	0.57073	0.13142	74	-0.65	001.99	0
001	Loss on Drying (%)	0089	8.6800	0.00000	9.0213	0.57073	0.13142	74	-0.60	001.07	0
001	Loss on Drying (%)	0644	8.6850	0.01000	9.0213	0.57073	0.13142	74	-0.59	001.05	0
001	Loss on Drying (%)	0940	8.6850	0.21000	9.0213	0.57073	0.13142	74	-0.59	001.99	0
001	Loss on Drying (%)	2145	8.7200	0.00000	9.0213	0.57073	0.13142	74	-0.53	001.07	0
001	Loss on Drying (%)	0186	8.7650	0.53000	9.0213	0.57073	0.13142	74	-0.45	001.03	0
001	Loss on Drying (%)	0620	8.7666	0.10570	9.0213	0.57073	0.13142	74	-0.45	001.00	0
001	Loss on Drying (%)	0903	8.7750	0.01000	9.0213	0.57073	0.13142	74	-0.43	001.03	0
001	Loss on Drying (%)	0171	8.7750	0.03000	9.0213	0.57073	0.13142	74	-0.43	001.07	0
001	Loss on Drying (%)	0893	8.8100	0.00000	9.0213	0.57073	0.13142	74	-0.37	001.03	0
001	Loss on Drying (%)	0571	8.8150	0.21000	9.0213	0.57073	0.13142	74	-0.36	001.07	0
001	Loss on Drying (%)	2146	8.8450	0.03000	9.0213	0.57073	0.13142	74	-0.31	001.07	0
001	Loss on Drying (%)	0619	8.8500	0.02000	9.0213	0.57073	0.13142	74	-0.30	001.03	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
001	Loss on Drying (%)	0413	8.8500	0.30000	9.0213	0.57073	0.13142	74	-0.30	001.07	0
001	Loss on Drying (%)	2059	8.8800	0.06000	9.0213	0.57073	0.13142	74	-0.25	001.07	0
001	Loss on Drying (%)	0003	8.9650	0.15000	9.0213	0.57073	0.13142	74	-0.10	001.07	0
001	Loss on Drying (%)	0074	9.0000	0.00000	9.0213	0.57073	0.13142	74	-0.04	001.07	0
001	Loss on Drying (%)	0035	9.0050	0.17000	9.0213	0.57073	0.13142	74	-0.03	001.07	0
001	Loss on Drying (%)	0683	9.0050	0.05000	9.0213	0.57073	0.13142	74	-0.03	001.07	0
001	Loss on Drying (%)	0843	9.0400	0.62000	9.0213	0.57073	0.13142	74	0.03	001.07	0
001	Loss on Drying (%)	2062	9.0440	0.11870	9.0213	0.57073	0.13142	74	0.04	001.03	0
001	Loss on Drying (%)	0612	9.0550	0.25000	9.0213	0.57073	0.13142	74	0.06	001.99	0
001	Loss on Drying (%)	0083	9.0750	0.03000	9.0213	0.57073	0.13142	74	0.09	001.07	0
001	Loss on Drying (%)	0590	9.0750	0.63000	9.0213	0.57073	0.13142	74	0.09	001.08	0
001	Loss on Drying (%)	2109	9.0950	0.13000	9.0213	0.57073	0.13142	74	0.13	001.07	0
001	Loss on Drying (%)	0629	9.1300	0.08000	9.0213	0.57073	0.13142	74	0.19	001.99	0
001	Loss on Drying (%)	2009	9.1302	0.01090	9.0213	0.57073	0.13142	74	0.19	001.07	0
001	Loss on Drying (%)	0187	9.1450	0.05000	9.0213	0.57073	0.13142	74	0.22	001.07	0
001	Loss on Drying (%)	0910	9.1500	0.10000	9.0213	0.57073	0.13142	74	0.23	001.99	0
001	Loss on Drying (%)	0226	9.1600	0.02000	9.0213	0.57073	0.13142	74	0.24	001.07	0
001	Loss on Drying (%)	0139	9.2200	0.02000	9.0213	0.57073	0.13142	74	0.35	001.07	0
001	Loss on Drying (%)	2053	9.2350	0.05000	9.0213	0.57073	0.13142	74	0.37	001.07	0
001	Loss on Drying (%)	0309	9.2480	0.21800	9.0213	0.57073	0.13142	74	0.40	001.00	0
001	Loss on Drying (%)	0652	9.3500	0.50000	9.0213	0.57073	0.13142	74	0.58	001.08	0
001	Loss on Drying (%)	0689	9.3500	0.10000	9.0213	0.57073	0.13142	74	0.58	001.07	0
001	Loss on Drying (%)	0693	9.3579	0.01940	9.0213	0.57073	0.13142	74	0.59	001.07	0
001	Loss on Drying (%)	0912	9.3900	0.26000	9.0213	0.57073	0.13142	74	0.65	001.99	0
001	Loss on Drying (%)	0142	9.3950	0.09000	9.0213	0.57073	0.13142	74	0.65	001.07	0
001	Loss on Drying (%)	0676	9.5500	0.02000	9.0213	0.57073	0.13142	74	0.93	001.99	0
001	Loss on Drying (%)	0849	9.5685	0.03700	9.0213	0.57073	0.13142	74	0.96	001.07	0
001	Loss on Drying (%)	0353	9.5950	0.47000	9.0213	0.57073	0.13142	74	1.01	001.07	0
001	Loss on Drying (%)	0045	9.6100	0.00000	9.0213	0.57073	0.13142	74	1.03	001.07	0
001	Loss on Drying (%)	0876	9.6150	0.25000	9.0213	0.57073	0.13142	74	1.04	001.99	0
001	Loss on Drying (%)	0357	9.6385	0.22940	9.0213	0.57073	0.13142	74	1.08	001.99	0
001	Loss on Drying (%)	0015	9.6650	0.03000	9.0213	0.57073	0.13142	74	1.13	001.07	0
001	Loss on Drying (%)	0098	9.7100	0.30000	9.0213	0.57073	0.13142	74	1.21	001.07	0
001	Loss on Drying (%)	0004	9.7150	0.07000	9.0213	0.57073	0.13142	74	1.22	001.99	0
001	Loss on Drying (%)	0505	9.7700	0.04000	9.0213	0.57073	0.13142	74	1.31	001.99	0
001	Loss on Drying (%)	0169	9.8850	0.01000	9.0213	0.57073	0.13142	74	1.51	001.00	0
001	Loss on Drying (%)	0581	9.9800	0.12000	9.0213	0.57073	0.13142	74	1.68	001.07	0
001	Loss on Drying (%)	0615	10.045	0.03000	9.0213	0.57073	0.13142	74	1.79	001.00	0
001	Loss on Drying (%)	0019	10.215	0.25000	9.0213	0.57073	0.13142	74	2.09	001.07	0
001	Loss on Drying (%)	0037	10.370	0.02000	9.0213	0.57073	0.13142	74	2.36	001.99	0
001	Loss on Drying (%)	0504	10.380	0.50000	9.0213	0.57073	0.13142	74	2.38	001.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
001	Loss on Drying (%)	0529	11.720	0.12000	9.0213	0.57073	0.13142	74	4.73	001.07	0
001	Loss on Drying (%)	0366	9.8000	0.80000	9.0213	0.57073	0.13142	74	1.36	001.07	1
001	Loss on Drying (%)	0872	21.285	0.05000	9.0213	0.57073	0.13142	74	21.49	001.07	2
002	Protein (%)	0539	19.775	0.33000	21.190	0.33028	0.18114	186	-4.28	002.06	0
002	Protein (%)	0874	20.095	0.27000	21.190	0.33028	0.18114	186	-3.32	002.04	0
002	Protein (%)	0685	20.155	0.01000	21.190	0.33028	0.18114	186	-3.13	002.01	0
002	Protein (%)	1019	20.221	0.10300	21.190	0.33028	0.18114	186	-2.94	002.06	0
002	Protein (%)	0062	20.400	0.20000	21.190	0.33028	0.18114	186	-2.39	002.01	0
002	Protein (%)	0028	20.420	0.04000	21.190	0.33028	0.18114	186	-2.33	002.01	0
002	Protein (%)	2075	20.430	0.38000	21.190	0.33028	0.18114	186	-2.30	002.06	0
002	Protein (%)	0042	20.455	0.13000	21.190	0.33028	0.18114	186	-2.23	002.02	0
002	Protein (%)	0407	20.517	0.41000	21.190	0.33028	0.18114	186	-2.04	002.06	0
002	Protein (%)	0623	20.538	0.39100	21.190	0.33028	0.18114	186	-1.98	002.05	0
002	Protein (%)	0889	20.550	0.86000	21.190	0.33028	0.18114	186	-1.94	002.01	0
002	Protein (%)	0689	20.550	0.50000	21.190	0.33028	0.18114	186	-1.94	002.05	0
002	Protein (%)	0676	20.565	0.25000	21.190	0.33028	0.18114	186	-1.89	002.06	0
002	Protein (%)	2012	20.590	0.08000	21.190	0.33028	0.18114	186	-1.82	002.05	0
002	Protein (%)	0653	20.600	0.18000	21.190	0.33028	0.18114	186	-1.79	002.01	0
002	Protein (%)	2012	20.700	0.26000	21.190	0.33028	0.18114	186	-1.48	002.06	0
002	Protein (%)	0652	20.700	0.20000	21.190	0.33028	0.18114	186	-1.48	002.01	0
002	Protein (%)	0592	20.715	0.09000	21.190	0.33028	0.18114	186	-1.44	002.06	0
002	Protein (%)	0042	20.750	0.10000	21.190	0.33028	0.18114	186	-1.33	002.06	0
002	Protein (%)	0955	20.750	0.10000	21.190	0.33028	0.18114	186	-1.33	002.06	0
002	Protein (%)	0563	20.770	0.06000	21.190	0.33028	0.18114	186	-1.27	002.08	0
002	Protein (%)	0921	20.780	0.00000	21.190	0.33028	0.18114	186	-1.24	002.06	0
002	Protein (%)	0619	20.800	0.20000	21.190	0.33028	0.18114	186	-1.18	002.05	0
002	Protein (%)	0692	20.800	0.60000	21.190	0.33028	0.18114	186	-1.18	002.06	0
002	Protein (%)	0620	20.810	0.16800	21.190	0.33028	0.18114	186	-1.15	002.05	0
002	Protein (%)	0675	20.815	0.03000	21.190	0.33028	0.18114	186	-1.14	002.05	0
002	Protein (%)	0870	20.824	0.08600	21.190	0.33028	0.18114	186	-1.11	002.01	0
002	Protein (%)	2145	20.835	0.23000	21.190	0.33028	0.18114	186	-1.08	002.05	0
002	Protein (%)	0683	20.850	0.04000	21.190	0.33028	0.18114	186	-1.03	002.05	0
002	Protein (%)	0294	20.850	0.10000	21.190	0.33028	0.18114	186	-1.03	002.06	0
002	Protein (%)	0043	20.860	0.08000	21.190	0.33028	0.18114	186	-1.00	002.01	0
002	Protein (%)	2146	20.860	0.06000	21.190	0.33028	0.18114	186	-1.00	002.05	0
002	Protein (%)	0885	20.865	0.02160	21.190	0.33028	0.18114	186	-0.98	002.05	0
002	Protein (%)	2081	20.865	0.63000	21.190	0.33028	0.18114	186	-0.98	002.06	0
002	Protein (%)	0169	20.875	0.03000	21.190	0.33028	0.18114	186	-0.95	002.00	0
002	Protein (%)	0505	20.890	0.02000	21.190	0.33028	0.18114	186	-0.91	002.06	0
002	Protein (%)	0955	20.900	0.00000	21.190	0.33028	0.18114	186	-0.88	002.05	0
002	Protein (%)	0226	20.900	0.20000	21.190	0.33028	0.18114	186	-0.88	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein (%)	0650	20.900	0.08000	21.190	0.33028	0.18114	186	-0.88	002.06	0
002	Protein (%)	0861	20.900	0.24000	21.190	0.33028	0.18114	186	-0.88	002.06	0
002	Protein (%)	0968	20.915	0.03400	21.190	0.33028	0.18114	186	-0.83	002.01	0
002	Protein (%)	0878	20.915	0.19000	21.190	0.33028	0.18114	186	-0.83	002.05	0
002	Protein (%)	0598	20.920	0.02000	21.190	0.33028	0.18114	186	-0.82	002.06	0
002	Protein (%)	2093	20.930	0.04000	21.190	0.33028	0.18114	186	-0.79	002.05	0
002	Protein (%)	2150	20.941	0.58300	21.190	0.33028	0.18114	186	-0.76	002.06	0
002	Protein (%)	0010	20.950	0.30000	21.190	0.33028	0.18114	186	-0.73	002.06	0
002	Protein (%)	0047	20.950	0.70000	21.190	0.33028	0.18114	186	-0.73	002.06	0
002	Protein (%)	0948	20.965	0.07000	21.190	0.33028	0.18114	186	-0.68	002.05	0
002	Protein (%)	0354	20.975	0.07000	21.190	0.33028	0.18114	186	-0.65	002.05	0
002	Protein (%)	0941	20.985	0.19000	21.190	0.33028	0.18114	186	-0.62	002.05	0
002	Protein (%)	0298	21.000	0.02000	21.190	0.33028	0.18114	186	-0.58	002.06	0
002	Protein (%)	0202	21.035	0.01000	21.190	0.33028	0.18114	186	-0.47	002.06	0
002	Protein (%)	0682	21.040	0.00000	21.190	0.33028	0.18114	186	-0.45	002.06	0
002	Protein (%)	2155	21.041	0.43200	21.190	0.33028	0.18114	186	-0.45	002.06	0
002	Protein (%)	0728	21.045	0.17000	21.190	0.33028	0.18114	186	-0.44	002.04	0
002	Protein (%)	0939	21.045	0.01000	21.190	0.33028	0.18114	186	-0.44	002.05	0
002	Protein (%)	2097	21.047	0.24660	21.190	0.33028	0.18114	186	-0.43	002.05	0
002	Protein (%)	0881	21.050	0.02700	21.190	0.33028	0.18114	186	-0.43	002.05	0
002	Protein (%)	0610	21.050	0.10000	21.190	0.33028	0.18114	186	-0.42	002.01	0
002	Protein (%)	0187	21.050	0.04000	21.190	0.33028	0.18114	186	-0.42	002.04	0
002	Protein (%)	2109	21.070	0.20000	21.190	0.33028	0.18114	186	-0.36	002.05	0
002	Protein (%)	0074	21.080	0.12000	21.190	0.33028	0.18114	186	-0.33	002.06	0
002	Protein (%)	0552	21.085	0.03000	21.190	0.33028	0.18114	186	-0.32	002.05	0
002	Protein (%)	0019	21.085	0.07000	21.190	0.33028	0.18114	186	-0.32	002.06	0
002	Protein (%)	0948	21.085	0.21000	21.190	0.33028	0.18114	186	-0.32	002.06	0
002	Protein (%)	0848	21.095	0.11000	21.190	0.33028	0.18114	186	-0.29	002.06	0
002	Protein (%)	0870	21.095	0.02500	21.190	0.33028	0.18114	186	-0.29	002.06	0
002	Protein (%)	0152	21.100	0.00000	21.190	0.33028	0.18114	186	-0.27	002.01	0
002	Protein (%)	0100	21.100	0.34000	21.190	0.33028	0.18114	186	-0.27	002.06	0
002	Protein (%)	2066	21.100	0.20000	21.190	0.33028	0.18114	186	-0.27	002.10	0
002	Protein (%)	0098	21.105	0.05000	21.190	0.33028	0.18114	186	-0.26	002.08	0
002	Protein (%)	0903	21.115	0.01000	21.190	0.33028	0.18114	186	-0.23	002.05	0
002	Protein (%)	2073	21.115	0.01000	21.190	0.33028	0.18114	186	-0.23	002.05	0
002	Protein (%)	0139	21.115	0.03000	21.190	0.33028	0.18114	186	-0.23	002.06	0
002	Protein (%)	0148	21.115	0.13000	21.190	0.33028	0.18114	186	-0.23	002.06	0
002	Protein (%)	0618	21.118	0.23800	21.190	0.33028	0.18114	186	-0.22	002.06	0
002	Protein (%)	0009	21.125	0.17000	21.190	0.33028	0.18114	186	-0.20	002.06	0
002	Protein (%)	0897	21.130	0.02000	21.190	0.33028	0.18114	186	-0.18	002.05	0
002	Protein (%)	0940	21.140	0.14000	21.190	0.33028	0.18114	186	-0.15	002.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein (%)	0581	21.140	0.66000	21.190	0.33028	0.18114	186	-0.15	002.04	0
002	Protein (%)	0895	21.140	0.02000	21.190	0.33028	0.18114	186	-0.15	002.05	0
002	Protein (%)	0520	21.140	0.12000	21.190	0.33028	0.18114	186	-0.15	002.06	0
002	Protein (%)	0034	21.146	0.48900	21.190	0.33028	0.18114	186	-0.14	002.06	0
002	Protein (%)	0952	21.150	0.02000	21.190	0.33028	0.18114	186	-0.12	002.05	0
002	Protein (%)	0610	21.150	0.10000	21.190	0.33028	0.18114	186	-0.12	002.06	0
002	Protein (%)	0004	21.155	0.17000	21.190	0.33028	0.18114	186	-0.11	002.06	0
002	Protein (%)	0049	21.155	0.33000	21.190	0.33028	0.18114	186	-0.11	002.06	0
002	Protein (%)	0508	21.159	0.69600	21.190	0.33028	0.18114	186	-0.09	002.06	0
002	Protein (%)	0144	21.170	0.02000	21.190	0.33028	0.18114	186	-0.06	002.06	0
002	Protein (%)	0849	21.183	0.05500	21.190	0.33028	0.18114	186	-0.02	002.05	0
002	Protein (%)	0629	21.195	0.05000	21.190	0.33028	0.18114	186	0.01	002.10	0
002	Protein (%)	0693	21.204	0.12620	21.190	0.33028	0.18114	186	0.04	002.06	0
002	Protein (%)	0142	21.205	0.05000	21.190	0.33028	0.18114	186	0.05	002.06	0
002	Protein (%)	0205	21.205	0.09000	21.190	0.33028	0.18114	186	0.05	002.06	0
002	Protein (%)	0571	21.210	0.04300	21.190	0.33028	0.18114	186	0.06	002.06	0
002	Protein (%)	0265	21.215	0.25000	21.190	0.33028	0.18114	186	0.08	002.06	0
002	Protein (%)	2129	21.215	0.33000	21.190	0.33028	0.18114	186	0.08	002.06	0
002	Protein (%)	2062	21.217	0.02540	21.190	0.33028	0.18114	186	0.08	002.05	0
002	Protein (%)	0651	21.229	0.14300	21.190	0.33028	0.18114	186	0.12	002.05	0
002	Protein (%)	2006	21.230	0.06000	21.190	0.33028	0.18114	186	0.12	002.05	0
002	Protein (%)	0546	21.235	0.03000	21.190	0.33028	0.18114	186	0.14	002.06	0
002	Protein (%)	0229	21.245	0.61000	21.190	0.33028	0.18114	186	0.17	002.06	0
002	Protein (%)	0036	21.245	0.13250	21.190	0.33028	0.18114	186	0.17	002.02	0
002	Protein (%)	0066	21.250	0.08000	21.190	0.33028	0.18114	186	0.18	002.02	0
002	Protein (%)	0357	21.250	0.25000	21.190	0.33028	0.18114	186	0.18	002.06	0
002	Protein (%)	0619	21.250	0.10000	21.190	0.33028	0.18114	186	0.18	002.06	0
002	Protein (%)	2057	21.264	0.30700	21.190	0.33028	0.18114	186	0.22	002.06	0
002	Protein (%)	0001	21.265	0.17000	21.190	0.33028	0.18114	186	0.23	002.06	0
002	Protein (%)	0843	21.265	0.23000	21.190	0.33028	0.18114	186	0.23	002.06	0
002	Protein (%)	0504	21.275	0.25000	21.190	0.33028	0.18114	186	0.26	002.04	0
002	Protein (%)	2059	21.277	0.13440	21.190	0.33028	0.18114	186	0.26	002.06	0
002	Protein (%)	0626	21.279	0.10400	21.190	0.33028	0.18114	186	0.27	002.06	0
002	Protein (%)	0164	21.290	0.06000	21.190	0.33028	0.18114	186	0.30	002.01	0
002	Protein (%)	2022	21.290	0.02000	21.190	0.33028	0.18114	186	0.30	002.05	0
002	Protein (%)	0026	21.291	0.27900	21.190	0.33028	0.18114	186	0.30	002.06	0
002	Protein (%)	0175	21.300	0.20000	21.190	0.33028	0.18114	186	0.33	002.06	0
002	Protein (%)	0297	21.300	0.02000	21.190	0.33028	0.18114	186	0.33	002.06	0
002	Protein (%)	0553	21.300	0.20000	21.190	0.33028	0.18114	186	0.33	002.06	0
002	Protein (%)	0366	21.300	0.00000	21.190	0.33028	0.18114	186	0.33	002.06	0
002	Protein (%)	0003	21.320	0.56000	21.190	0.33028	0.18114	186	0.39	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein (%)	2009	21.333	0.08610	21.190	0.33028	0.18114	186	0.43	002.05	0
002	Protein (%)	0083	21.345	0.15000	21.190	0.33028	0.18114	186	0.47	002.06	0
002	Protein (%)	0033	21.350	0.30000	21.190	0.33028	0.18114	186	0.48	002.06	0
002	Protein (%)	0278	21.350	0.10000	21.190	0.33028	0.18114	186	0.48	002.06	0
002	Protein (%)	0413	21.350	0.30000	21.190	0.33028	0.18114	186	0.48	002.06	0
002	Protein (%)	0510	21.350	0.10000	21.190	0.33028	0.18114	186	0.48	002.06	0
002	Protein (%)	0910	21.350	0.30000	21.190	0.33028	0.18114	186	0.48	002.06	0
002	Protein (%)	0536	21.370	0.34000	21.190	0.33028	0.18114	186	0.54	002.05	0
002	Protein (%)	0590	21.375	0.15000	21.190	0.33028	0.18114	186	0.56	002.06	0
002	Protein (%)	0504	21.380	0.12000	21.190	0.33028	0.18114	186	0.57	002.06	0
002	Protein (%)	0675	21.380	0.12000	21.190	0.33028	0.18114	186	0.57	002.06	0
002	Protein (%)	0857	21.380	0.30000	21.190	0.33028	0.18114	186	0.57	002.06	0
002	Protein (%)	0622	21.385	0.36000	21.190	0.33028	0.18114	186	0.59	002.05	0
002	Protein (%)	0683	21.395	0.27000	21.190	0.33028	0.18114	186	0.62	002.06	0
002	Protein (%)	0045	21.400	0.20000	21.190	0.33028	0.18114	186	0.64	002.06	0
002	Protein (%)	0171	21.400	0.20000	21.190	0.33028	0.18114	186	0.64	002.06	0
002	Protein (%)	0353	21.400	0.08000	21.190	0.33028	0.18114	186	0.64	002.06	0
002	Protein (%)	0208	21.410	0.20000	21.190	0.33028	0.18114	186	0.67	002.06	0
002	Protein (%)	0358	21.410	0.00000	21.190	0.33028	0.18114	186	0.67	002.06	0
002	Protein (%)	0630	21.420	0.12000	21.190	0.33028	0.18114	186	0.70	002.06	0
002	Protein (%)	0354	21.430	0.02000	21.190	0.33028	0.18114	186	0.73	002.06	0
002	Protein (%)	0164	21.435	0.01000	21.190	0.33028	0.18114	186	0.74	002.06	0
002	Protein (%)	0425	21.445	0.05000	21.190	0.33028	0.18114	186	0.77	002.06	0
002	Protein (%)	0014	21.450	0.10000	21.190	0.33028	0.18114	186	0.79	002.06	0
002	Protein (%)	0108	21.450	0.22000	21.190	0.33028	0.18114	186	0.79	002.06	0
002	Protein (%)	0687	21.450	0.10000	21.190	0.33028	0.18114	186	0.79	002.06	0
002	Protein (%)	2004	21.450	0.30000	21.190	0.33028	0.18114	186	0.79	002.99	0
002	Protein (%)	0541	21.455	0.21000	21.190	0.33028	0.18114	186	0.80	002.06	0
002	Protein (%)	0889	21.460	0.30000	21.190	0.33028	0.18114	186	0.82	002.11	0
002	Protein (%)	0194	21.465	0.03000	21.190	0.33028	0.18114	186	0.83	002.05	0
002	Protein (%)	0036	21.465	0.03000	21.190	0.33028	0.18114	186	0.83	002.06	0
002	Protein (%)	0574	21.465	0.17000	21.190	0.33028	0.18114	186	0.83	002.06	0
002	Protein (%)	2053	21.480	0.26000	21.190	0.33028	0.18114	186	0.88	002.06	0
002	Protein (%)	0309	21.483	0.11000	21.190	0.33028	0.18114	186	0.89	002.06	0
002	Protein (%)	0964	21.485	0.27000	21.190	0.33028	0.18114	186	0.89	002.06	0
002	Protein (%)	0035	21.495	0.13000	21.190	0.33028	0.18114	186	0.92	002.06	0
002	Protein (%)	0098	21.500	0.20000	21.190	0.33028	0.18114	186	0.94	002.06	0
002	Protein (%)	0263	21.505	0.10870	21.190	0.33028	0.18114	186	0.95	002.06	0
002	Protein (%)	0939	21.505	0.07000	21.190	0.33028	0.18114	186	0.95	002.06	0
002	Protein (%)	0043	21.510	0.20000	21.190	0.33028	0.18114	186	0.97	002.06	0
002	Protein (%)	2152	21.520	0.04000	21.190	0.33028	0.18114	186	1.00	002.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein (%)	0233	21.525	0.35000	21.190	0.33028	0.18114	186	1.01	002.06	0
002	Protein (%)	0037	21.540	0.42000	21.190	0.33028	0.18114	186	1.06	002.06	0
002	Protein (%)	0529	21.540	0.02000	21.190	0.33028	0.18114	186	1.06	002.06	0
002	Protein (%)	2006	21.550	0.20000	21.190	0.33028	0.18114	186	1.09	002.06	0
002	Protein (%)	0300	21.551	0.27800	21.190	0.33028	0.18114	186	1.09	002.06	0
002	Protein (%)	0644	21.556	0.70500	21.190	0.33028	0.18114	186	1.11	002.06	0
002	Protein (%)	0904	21.565	0.07000	21.190	0.33028	0.18114	186	1.14	002.06	0
002	Protein (%)	0859	21.569	0.12400	21.190	0.33028	0.18114	186	1.15	002.06	0
002	Protein (%)	0190	21.600	0.12000	21.190	0.33028	0.18114	186	1.24	002.06	0
002	Protein (%)	0726	21.625	0.01000	21.190	0.33028	0.18114	186	1.32	002.06	0
002	Protein (%)	0511	21.635	0.29000	21.190	0.33028	0.18114	186	1.35	002.06	0
002	Protein (%)	0956	21.650	0.30000	21.190	0.33028	0.18114	186	1.39	002.06	0
002	Protein (%)	0186	21.700	0.20000	21.190	0.33028	0.18114	186	1.54	002.06	0
002	Protein (%)	0941	21.765	0.45000	21.190	0.33028	0.18114	186	1.74	002.06	0
002	Protein (%)	0589	21.770	0.22000	21.190	0.33028	0.18114	186	1.76	002.06	0
002	Protein (%)	0893	21.780	0.00000	21.190	0.33028	0.18114	186	1.79	002.05	0
002	Protein (%)	0297	21.800	0.20000	21.190	0.33028	0.18114	186	1.85	002.11	0
002	Protein (%)	2089	21.860	0.00000	21.190	0.33028	0.18114	186	2.03	002.06	0
002	Protein (%)	0953	22.035	0.67000	21.190	0.33028	0.18114	186	2.56	002.06	0
002	Protein (%)	0615	22.200	0.40000	21.190	0.33028	0.18114	186	3.06	002.06	0
002	Protein (%)	0345	22.250	0.30000	21.190	0.33028	0.18114	186	3.21	002.06	0
002	Protein (%)	0612	23.115	0.77000	21.190	0.33028	0.18114	186	5.83	002.06	0
002	Protein (%)	2105	23.210	0.08000	21.190	0.33028	0.18114	186	6.12	002.99	0
002	Protein (%)	0876	23.325	0.07000	21.190	0.33028	0.18114	186	6.46	002.06	0
002	Protein (%)	0405	23.490	0.02000	21.190	0.33028	0.18114	186	6.96	002.04	0
002	Protein (%)	0910	20.100	1.4000	21.190	0.33028	0.18114	186	-3.30	002.01	1
002	Protein (%)	0918	21.305	0.91000	21.190	0.33028	0.18114	186	0.35	002.06	1
002	Protein (%)	0872	9.4250	0.01000	21.190	0.33028	0.18114	186	-35.62	002.06	2
002	Protein (%)	0183	32.200	0.04000	21.190	0.33028	0.18114	186	33.33	002.01	2
003	Fat (%)	0848	5.6300	0.12000	6.6400	0.31814	0.11046	142	-3.17	003.14	0
003	Fat (%)	0504	5.6600	0.28000	6.6400	0.31814	0.11046	142	-3.08	003.01	0
003	Fat (%)	0546	5.6600	0.50000	6.6400	0.31814	0.11046	142	-3.08	003.99	0
003	Fat (%)	0881	5.7720	0.02400	6.6400	0.31814	0.11046	142	-2.73	003.10	0
003	Fat (%)	2022	5.9550	0.07000	6.6400	0.31814	0.11046	142	-2.15	003.10	0
003	Fat (%)	0619	5.9600	0.12000	6.6400	0.31814	0.11046	142	-2.14	003.10	0
003	Fat (%)	0019	6.0900	0.20000	6.6400	0.31814	0.11046	142	-1.73	003.14	0
003	Fat (%)	2068	6.1150	0.23000	6.6400	0.31814	0.11046	142	-1.65	003.09	0
003	Fat (%)	0685	6.1350	0.01000	6.6400	0.31814	0.11046	142	-1.59	003.09	0
003	Fat (%)	0169	6.1800	0.00000	6.6400	0.31814	0.11046	142	-1.45	003.06	0
003	Fat (%)	0298	6.1800	0.02000	6.6400	0.31814	0.11046	142	-1.45	003.10	0
003	Fat (%)	2073	6.1950	0.21000	6.6400	0.31814	0.11046	142	-1.40	003.10	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat (%)	0581	6.1950	0.13000	6.6400	0.31814	0.11046	142	-1.40	003.14	0
003	Fat (%)	0918	6.2500	0.04000	6.6400	0.31814	0.11046	142	-1.23	003.06	0
003	Fat (%)	0693	6.2514	0.08060	6.6400	0.31814	0.11046	142	-1.22	003.10	0
003	Fat (%)	0682	6.2600	0.00000	6.6400	0.31814	0.11046	142	-1.19	003.14	0
003	Fat (%)	0062	6.3000	0.00000	6.6400	0.31814	0.11046	142	-1.07	003.10	0
003	Fat (%)	2152	6.3000	0.08000	6.6400	0.31814	0.11046	142	-1.07	003.12	0
003	Fat (%)	0574	6.3000	0.04000	6.6400	0.31814	0.11046	142	-1.07	003.14	0
003	Fat (%)	0675	6.3150	0.13000	6.6400	0.31814	0.11046	142	-1.02	003.09	0
003	Fat (%)	0895	6.3350	0.03000	6.6400	0.31814	0.11046	142	-0.96	003.12	0
003	Fat (%)	2009	6.3357	0.02300	6.6400	0.31814	0.11046	142	-0.96	003.10	0
003	Fat (%)	2012	6.3400	0.06000	6.6400	0.31814	0.11046	142	-0.94	003.09	0
003	Fat (%)	0175	6.3400	0.20000	6.6400	0.31814	0.11046	142	-0.94	003.14	0
003	Fat (%)	0728	6.3450	0.01000	6.6400	0.31814	0.11046	142	-0.93	003.10	0
003	Fat (%)	0651	6.3480	0.01200	6.6400	0.31814	0.11046	142	-0.92	003.10	0
003	Fat (%)	0187	6.3600	0.30000	6.6400	0.31814	0.11046	142	-0.88	003.13	0
003	Fat (%)	0354	6.3650	0.01000	6.6400	0.31814	0.11046	142	-0.86	003.09	0
003	Fat (%)	0353	6.3850	0.09000	6.6400	0.31814	0.11046	142	-0.80	003.09	0
003	Fat (%)	0897	6.3850	0.01000	6.6400	0.31814	0.11046	142	-0.80	003.10	0
003	Fat (%)	0520	6.3850	0.41000	6.6400	0.31814	0.11046	142	-0.80	003.14	0
003	Fat (%)	0226	6.4000	0.20000	6.6400	0.31814	0.11046	142	-0.75	003.09	0
003	Fat (%)	0042	6.4000	0.32000	6.6400	0.31814	0.11046	142	-0.75	003.10	0
003	Fat (%)	0366	6.4000	0.20000	6.6400	0.31814	0.11046	142	-0.75	003.10	0
003	Fat (%)	0629	6.4000	0.14000	6.6400	0.31814	0.11046	142	-0.75	003.10	0
003	Fat (%)	0108	6.4000	0.32000	6.6400	0.31814	0.11046	142	-0.75	003.14	0
003	Fat (%)	0100	6.4050	0.11000	6.6400	0.31814	0.11046	142	-0.74	003.10	0
003	Fat (%)	0644	6.4150	0.07000	6.6400	0.31814	0.11046	142	-0.71	003.06	0
003	Fat (%)	0541	6.4150	0.33000	6.6400	0.31814	0.11046	142	-0.71	003.10	0
003	Fat (%)	0037	6.4200	0.04000	6.6400	0.31814	0.11046	142	-0.69	003.14	0
003	Fat (%)	0208	6.4350	0.03000	6.6400	0.31814	0.11046	142	-0.64	003.13	0
003	Fat (%)	0152	6.4500	0.10000	6.6400	0.31814	0.11046	142	-0.60	003.09	0
003	Fat (%)	0508	6.4507	0.00460	6.6400	0.31814	0.11046	142	-0.60	003.09	0
003	Fat (%)	0653	6.4600	0.04000	6.6400	0.31814	0.11046	142	-0.57	003.00	0
003	Fat (%)	0553	6.4700	0.12000	6.6400	0.31814	0.11046	142	-0.53	003.10	0
003	Fat (%)	0687	6.4750	0.15000	6.6400	0.31814	0.11046	142	-0.52	003.06	0
003	Fat (%)	0903	6.4750	0.01000	6.6400	0.31814	0.11046	142	-0.52	003.12	0
003	Fat (%)	0956	6.5000	0.00000	6.6400	0.31814	0.11046	142	-0.44	003.14	0
003	Fat (%)	0425	6.5050	0.05000	6.6400	0.31814	0.11046	142	-0.42	003.06	0
003	Fat (%)	0164	6.5100	0.06000	6.6400	0.31814	0.11046	142	-0.41	003.06	0
003	Fat (%)	0968	6.5205	0.01300	6.6400	0.31814	0.11046	142	-0.38	003.14	0
003	Fat (%)	0912	6.5250	0.07000	6.6400	0.31814	0.11046	142	-0.36	003.14	0
003	Fat (%)	0294	6.5400	0.10000	6.6400	0.31814	0.11046	142	-0.31	003.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat (%)	0357	6.5500	0.10000	6.6400	0.31814	0.11046	142	-0.28	003.14	0
003	Fat (%)	2062	6.5538	0.02720	6.6400	0.31814	0.11046	142	-0.27	003.12	0
003	Fat (%)	0563	6.5701	0.27830	6.6400	0.31814	0.11046	142	-0.22	003.06	0
003	Fat (%)	2109	6.5750	0.13000	6.6400	0.31814	0.11046	142	-0.20	003.00	0
003	Fat (%)	0229	6.5750	0.03000	6.6400	0.31814	0.11046	142	-0.20	003.14	0
003	Fat (%)	0148	6.5800	0.06000	6.6400	0.31814	0.11046	142	-0.19	003.06	0
003	Fat (%)	2129	6.5850	0.03000	6.6400	0.31814	0.11046	142	-0.17	003.14	0
003	Fat (%)	0940	6.5950	0.09000	6.6400	0.31814	0.11046	142	-0.14	003.12	0
003	Fat (%)	0689	6.6000	0.20000	6.6400	0.31814	0.11046	142	-0.13	003.06	0
003	Fat (%)	0683	6.6000	0.06000	6.6400	0.31814	0.11046	142	-0.13	003.14	0
003	Fat (%)	0692	6.6000	0.20000	6.6400	0.31814	0.11046	142	-0.13	003.14	0
003	Fat (%)	0142	6.6050	0.05000	6.6400	0.31814	0.11046	142	-0.11	003.00	0
003	Fat (%)	0529	6.6050	0.15000	6.6400	0.31814	0.11046	142	-0.11	003.14	0
003	Fat (%)	0876	6.6050	0.03000	6.6400	0.31814	0.11046	142	-0.11	003.14	0
003	Fat (%)	0089	6.6100	0.00000	6.6400	0.31814	0.11046	142	-0.09	003.14	0
003	Fat (%)	0144	6.6200	0.04000	6.6400	0.31814	0.11046	142	-0.06	003.14	0
003	Fat (%)	0870	6.6234	0.35840	6.6400	0.31814	0.11046	142	-0.05	003.10	0
003	Fat (%)	0358	6.6350	0.09000	6.6400	0.31814	0.11046	142	-0.02	003.14	0
003	Fat (%)	0026	6.6450	0.11000	6.6400	0.31814	0.11046	142	0.02	003.00	0
003	Fat (%)	0552	6.6450	0.05000	6.6400	0.31814	0.11046	142	0.02	003.06	0
003	Fat (%)	0049	6.6450	0.45000	6.6400	0.31814	0.11046	142	0.02	003.14	0
003	Fat (%)	0033	6.6500	0.12000	6.6400	0.31814	0.11046	142	0.03	003.13	0
003	Fat (%)	0036	6.6505	0.02770	6.6400	0.31814	0.11046	142	0.03	003.10	0
003	Fat (%)	0175	6.6600	0.10000	6.6400	0.31814	0.11046	142	0.06	003.00	0
003	Fat (%)	2155	6.6785	0.14900	6.6400	0.31814	0.11046	142	0.12	003.10	0
003	Fat (%)	0034	6.6900	0.44000	6.6400	0.31814	0.11046	142	0.16	003.14	0
003	Fat (%)	0948	6.6950	0.11000	6.6400	0.31814	0.11046	142	0.17	003.09	0
003	Fat (%)	0278	6.7000	0.20000	6.6400	0.31814	0.11046	142	0.19	003.14	0
003	Fat (%)	0202	6.7000	0.06000	6.6400	0.31814	0.11046	142	0.19	003.14	0
003	Fat (%)	0233	6.7050	0.27000	6.6400	0.31814	0.11046	142	0.20	003.10	0
003	Fat (%)	0066	6.7050	0.29000	6.6400	0.31814	0.11046	142	0.20	003.14	0
003	Fat (%)	0171	6.7100	0.12000	6.6400	0.31814	0.11046	142	0.22	003.12	0
003	Fat (%)	0297	6.7250	0.11000	6.6400	0.31814	0.11046	142	0.27	003.06	0
003	Fat (%)	0263	6.7449	0.00340	6.6400	0.31814	0.11046	142	0.33	003.09	0
003	Fat (%)	0205	6.7450	0.07000	6.6400	0.31814	0.11046	142	0.33	003.13	0
003	Fat (%)	0536	6.7450	0.03000	6.6400	0.31814	0.11046	142	0.33	003.99	0
003	Fat (%)	0689	6.7500	0.10000	6.6400	0.31814	0.11046	142	0.35	003.10	0
003	Fat (%)	0083	6.7650	0.25000	6.6400	0.31814	0.11046	142	0.39	003.06	0
003	Fat (%)	0893	6.7700	0.00000	6.6400	0.31814	0.11046	142	0.41	003.12	0
003	Fat (%)	0849	6.7780	0.01600	6.6400	0.31814	0.11046	142	0.43	003.06	0
003	Fat (%)	0861	6.7800	0.02000	6.6400	0.31814	0.11046	142	0.44	003.10	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat (%)	0098	6.7850	0.05000	6.6400	0.31814	0.11046	142	0.46	003.10	0
003	Fat (%)	0186	6.7900	0.04000	6.6400	0.31814	0.11046	142	0.47	003.14	0
003	Fat (%)	0889	6.7950	0.09000	6.6400	0.31814	0.11046	142	0.49	003.10	0
003	Fat (%)	0511	6.8050	0.03000	6.6400	0.31814	0.11046	142	0.52	003.06	0
003	Fat (%)	0939	6.8100	0.12000	6.6400	0.31814	0.11046	142	0.53	003.00	0
003	Fat (%)	0878	6.8100	0.14000	6.6400	0.31814	0.11046	142	0.53	003.14	0
003	Fat (%)	0229	6.8200	0.08000	6.6400	0.31814	0.11046	142	0.57	003.06	0
003	Fat (%)	0098	6.8350	0.01000	6.6400	0.31814	0.11046	142	0.61	003.09	0
003	Fat (%)	0345	6.8500	0.10000	6.6400	0.31814	0.11046	142	0.66	003.00	0
003	Fat (%)	0045	6.8500	0.08000	6.6400	0.31814	0.11046	142	0.66	003.10	0
003	Fat (%)	0618	6.8530	0.06000	6.6400	0.31814	0.11046	142	0.67	003.10	0
003	Fat (%)	0619	6.8550	0.15000	6.6400	0.31814	0.11046	142	0.68	003.14	0
003	Fat (%)	0948	6.8550	0.05000	6.6400	0.31814	0.11046	142	0.68	003.14	0
003	Fat (%)	2093	6.8600	0.04000	6.6400	0.31814	0.11046	142	0.69	003.10	0
003	Fat (%)	0028	6.8700	0.02000	6.6400	0.31814	0.11046	142	0.72	003.12	0
003	Fat (%)	0004	6.8700	0.32000	6.6400	0.31814	0.11046	142	0.72	003.14	0
003	Fat (%)	0407	6.8794	0.01370	6.6400	0.31814	0.11046	142	0.75	003.14	0
003	Fat (%)	0004	6.8800	0.18000	6.6400	0.31814	0.11046	142	0.75	003.09	0
003	Fat (%)	0955	6.8850	0.01000	6.6400	0.31814	0.11046	142	0.77	003.14	0
003	Fat (%)	0505	6.8900	0.06000	6.6400	0.31814	0.11046	142	0.79	003.09	0
003	Fat (%)	0001	6.8965	0.11500	6.6400	0.31814	0.11046	142	0.81	003.14	0
003	Fat (%)	0682	6.9000	0.00000	6.6400	0.31814	0.11046	142	0.82	003.06	0
003	Fat (%)	0910	6.9000	0.04000	6.6400	0.31814	0.11046	142	0.82	003.06	0
003	Fat (%)	0098	6.9000	0.14000	6.6400	0.31814	0.11046	142	0.82	003.13	0
003	Fat (%)	0885	6.9052	0.02670	6.6400	0.31814	0.11046	142	0.83	003.00	0
003	Fat (%)	0300	6.9150	0.07000	6.6400	0.31814	0.11046	142	0.86	003.06	0
003	Fat (%)	0009	6.9150	0.13000	6.6400	0.31814	0.11046	142	0.86	003.14	0
003	Fat (%)	0590	6.9550	0.11000	6.6400	0.31814	0.11046	142	0.99	003.09	0
003	Fat (%)	0190	6.9750	0.19000	6.6400	0.31814	0.11046	142	1.05	003.14	0
003	Fat (%)	0265	6.9850	0.03000	6.6400	0.31814	0.11046	142	1.08	003.14	0
003	Fat (%)	0630	7.0150	0.07000	6.6400	0.31814	0.11046	142	1.18	003.09	0
003	Fat (%)	0615	7.0400	0.10000	6.6400	0.31814	0.11046	142	1.26	003.00	0
003	Fat (%)	0510	7.0500	0.10000	6.6400	0.31814	0.11046	142	1.29	003.09	0
003	Fat (%)	0309	7.0561	0.04260	6.6400	0.31814	0.11046	142	1.31	003.00	0
003	Fat (%)	0074	7.0800	0.08000	6.6400	0.31814	0.11046	142	1.38	003.14	0
003	Fat (%)	2053	7.1000	0.02000	6.6400	0.31814	0.11046	142	1.45	003.00	0
003	Fat (%)	2066	7.1100	0.16000	6.6400	0.31814	0.11046	142	1.48	003.06	0
003	Fat (%)	2006	7.1100	0.08000	6.6400	0.31814	0.11046	142	1.48	003.13	0
003	Fat (%)	0035	7.1800	0.08000	6.6400	0.31814	0.11046	142	1.70	003.00	0
003	Fat (%)	0964	7.2000	0.56000	6.6400	0.31814	0.11046	142	1.76	003.09	0
003	Fat (%)	0194	7.2050	0.03000	6.6400	0.31814	0.11046	142	1.78	003.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat (%)	0726	7.2050	0.03000	6.6400	0.31814	0.11046	142	1.78	003.14	0
003	Fat (%)	0413	7.4000	0.00000	6.6400	0.31814	0.11046	142	2.39	003.14	0
003	Fat (%)	0941	7.4850	0.21000	6.6400	0.31814	0.11046	142	2.66	003.99	0
003	Fat (%)	2075	7.5500	0.32000	6.6400	0.31814	0.11046	142	2.86	003.14	0
003	Fat (%)	0889	7.7750	0.11000	6.6400	0.31814	0.11046	142	3.57	003.11	0
003	Fat (%)	0297	8.1200	0.10000	6.6400	0.31814	0.11046	142	4.65	003.11	0
003	Fat (%)	1019	9.3600	0.16000	6.6400	0.31814	0.11046	142	8.55	003.99	0
003	Fat (%)	0843	6.1950	0.65000	6.6400	0.31814	0.11046	142	-1.40	003.14	1
003	Fat (%)	2089	6.6000	0.58000	6.6400	0.31814	0.11046	142	-0.13	003.13	1
003	Fat (%)	0623	7.3800	0.64800	6.6400	0.31814	0.11046	142	2.33	003.10	1
003	Fat (%)	0952	1.8100	0.02000	6.6400	0.31814	0.11046	142	-15.18	003.10	2
004	Fiber (%)	2009	0.56285	0.04250	1.8257	0.23252	0.10910	118	-5.43	004.07	0
004	Fiber (%)	0265	0.76000	0.12000	1.8257	0.23252	0.10910	118	-4.58	004.07	0
004	Fiber (%)	0144	1.0750	0.01000	1.8257	0.23252	0.10910	118	-3.23	004.07	0
004	Fiber (%)	0202	1.0900	0.04000	1.8257	0.23252	0.10910	118	-3.16	004.07	0
004	Fiber (%)	0629	1.3000	0.00000	1.8257	0.23252	0.10910	118	-2.26	004.99	0
004	Fiber (%)	0552	1.3400	0.08000	1.8257	0.23252	0.10910	118	-2.09	004.07	0
004	Fiber (%)	2012	1.3550	0.13000	1.8257	0.23252	0.10910	118	-2.02	004.07	0
004	Fiber (%)	0510	1.4000	0.20000	1.8257	0.23252	0.10910	118	-1.83	004.07	0
004	Fiber (%)	0683	1.4300	0.04000	1.8257	0.23252	0.10910	118	-1.70	004.07	0
004	Fiber (%)	0294	1.4500	0.10000	1.8257	0.23252	0.10910	118	-1.62	004.07	0
004	Fiber (%)	0910	1.4500	0.10000	1.8257	0.23252	0.10910	118	-1.62	004.07	0
004	Fiber (%)	0939	1.4850	0.09000	1.8257	0.23252	0.10910	118	-1.47	004.07	0
004	Fiber (%)	0169	1.4900	0.06000	1.8257	0.23252	0.10910	118	-1.44	004.00	0
004	Fiber (%)	0035	1.5250	0.11000	1.8257	0.23252	0.10910	118	-1.29	004.07	0
004	Fiber (%)	2089	1.5300	0.22000	1.8257	0.23252	0.10910	118	-1.27	004.03	0
004	Fiber (%)	0921	1.5350	0.01000	1.8257	0.23252	0.10910	118	-1.25	004.03	0
004	Fiber (%)	0726	1.5500	0.10000	1.8257	0.23252	0.10910	118	-1.19	004.07	0
004	Fiber (%)	0653	1.5750	0.03000	1.8257	0.23252	0.10910	118	-1.08	004.07	0
004	Fiber (%)	2146	1.5800	0.08000	1.8257	0.23252	0.10910	118	-1.06	004.07	0
004	Fiber (%)	0026	1.6000	0.00000	1.8257	0.23252	0.10910	118	-0.97	004.07	0
004	Fiber (%)	0610	1.6000	0.00000	1.8257	0.23252	0.10910	118	-0.97	004.99	0
004	Fiber (%)	0893	1.6400	0.00000	1.8257	0.23252	0.10910	118	-0.80	004.06	0
004	Fiber (%)	0229	1.6500	0.10000	1.8257	0.23252	0.10910	118	-0.76	004.07	0
004	Fiber (%)	0278	1.6500	0.30000	1.8257	0.23252	0.10910	118	-0.76	004.07	0
004	Fiber (%)	0413	1.6500	0.10000	1.8257	0.23252	0.10910	118	-0.76	004.07	0
004	Fiber (%)	0042	1.6550	0.45000	1.8257	0.23252	0.10910	118	-0.73	004.07	0
004	Fiber (%)	2053	1.6900	0.02000	1.8257	0.23252	0.10910	118	-0.58	004.07	0
004	Fiber (%)	0849	1.6945	0.01100	1.8257	0.23252	0.10910	118	-0.56	004.07	0
004	Fiber (%)	0066	1.6950	0.15000	1.8257	0.23252	0.10910	118	-0.56	004.07	0
004	Fiber (%)	0885	1.6996	0.02530	1.8257	0.23252	0.10910	118	-0.54	004.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber (%)	2145	1.7000	0.06000	1.8257	0.23252	0.10910	118	-0.54	004.06	0
004	Fiber (%)	0968	1.7050	0.15000	1.8257	0.23252	0.10910	118	-0.52	004.07	0
004	Fiber (%)	0009	1.7150	0.15000	1.8257	0.23252	0.10910	118	-0.48	004.07	0
004	Fiber (%)	0529	1.7300	0.08000	1.8257	0.23252	0.10910	118	-0.41	004.07	0
004	Fiber (%)	0948	1.7350	0.03000	1.8257	0.23252	0.10910	118	-0.39	004.06	0
004	Fiber (%)	0897	1.7400	0.02000	1.8257	0.23252	0.10910	118	-0.37	004.06	0
004	Fiber (%)	0034	1.7400	0.26000	1.8257	0.23252	0.10910	118	-0.37	004.07	0
004	Fiber (%)	0190	1.7500	0.10000	1.8257	0.23252	0.10910	118	-0.33	004.07	0
004	Fiber (%)	0708	1.7500	0.10000	1.8257	0.23252	0.10910	118	-0.33	004.07	0
004	Fiber (%)	0956	1.7500	0.10000	1.8257	0.23252	0.10910	118	-0.33	004.07	0
004	Fiber (%)	0354	1.7600	0.00000	1.8257	0.23252	0.10910	118	-0.28	004.00	0
004	Fiber (%)	0619	1.7675	0.08300	1.8257	0.23252	0.10910	118	-0.25	004.07	0
004	Fiber (%)	2066	1.7750	0.07000	1.8257	0.23252	0.10910	118	-0.22	004.00	0
004	Fiber (%)	0083	1.7750	0.01000	1.8257	0.23252	0.10910	118	-0.22	004.07	0
004	Fiber (%)	0870	1.7798	0.03800	1.8257	0.23252	0.10910	118	-0.20	004.07	0
004	Fiber (%)	0100	1.7800	0.12000	1.8257	0.23252	0.10910	118	-0.20	004.07	0
004	Fiber (%)	0263	1.7899	0.00660	1.8257	0.23252	0.10910	118	-0.15	004.07	0
004	Fiber (%)	0940	1.7900	0.30000	1.8257	0.23252	0.10910	118	-0.15	004.99	0
004	Fiber (%)	2068	1.7950	0.01000	1.8257	0.23252	0.10910	118	-0.13	004.06	0
004	Fiber (%)	0033	1.7950	0.45000	1.8257	0.23252	0.10910	118	-0.13	004.07	0
004	Fiber (%)	0948	1.7950	0.05000	1.8257	0.23252	0.10910	118	-0.13	004.07	0
004	Fiber (%)	0581	1.8000	0.20000	1.8257	0.23252	0.10910	118	-0.11	004.07	0
004	Fiber (%)	0689	1.8000	0.00000	1.8257	0.23252	0.10910	118	-0.11	004.06	0
004	Fiber (%)	2075	1.8000	0.00000	1.8257	0.23252	0.10910	118	-0.11	004.07	0
004	Fiber (%)	0353	1.8050	0.03000	1.8257	0.23252	0.10910	118	-0.09	004.03	0
004	Fiber (%)	0171	1.8100	0.08000	1.8257	0.23252	0.10910	118	-0.07	004.00	0
004	Fiber (%)	0941	1.8150	0.23000	1.8257	0.23252	0.10910	118	-0.05	004.06	0
004	Fiber (%)	0504	1.8250	0.13000	1.8257	0.23252	0.10910	118	0.00	004.00	0
004	Fiber (%)	0590	1.8250	0.15000	1.8257	0.23252	0.10910	118	0.00	004.06	0
004	Fiber (%)	0881	1.8325	0.13900	1.8257	0.23252	0.10910	118	0.03	004.06	0
004	Fiber (%)	0968	1.8345	0.02300	1.8257	0.23252	0.10910	118	0.04	004.06	0
004	Fiber (%)	0194	1.8350	0.03000	1.8257	0.23252	0.10910	118	0.04	004.00	0
004	Fiber (%)	0019	1.8400	0.42000	1.8257	0.23252	0.10910	118	0.06	004.07	0
004	Fiber (%)	0861	1.8400	0.10000	1.8257	0.23252	0.10910	118	0.06	004.07	0
004	Fiber (%)	0098	1.8400	0.02000	1.8257	0.23252	0.10910	118	0.06	004.06	0
004	Fiber (%)	0598	1.8450	0.11000	1.8257	0.23252	0.10910	118	0.08	004.07	0
004	Fiber (%)	0164	1.8500	0.10000	1.8257	0.23252	0.10910	118	0.10	004.00	0
004	Fiber (%)	0098	1.8500	0.08000	1.8257	0.23252	0.10910	118	0.10	004.07	0
004	Fiber (%)	0692	1.8500	0.30000	1.8257	0.23252	0.10910	118	0.10	004.07	0
004	Fiber (%)	0693	1.8555	0.15850	1.8257	0.23252	0.10910	118	0.13	004.03	0
004	Fiber (%)	0297	1.8600	0.18000	1.8257	0.23252	0.10910	118	0.15	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber (%)	0003	1.8700	0.24000	1.8257	0.23252	0.10910	118	0.19	004.07	0
004	Fiber (%)	0309	1.8770	0.06840	1.8257	0.23252	0.10910	118	0.22	004.00	0
004	Fiber (%)	2073	1.8800	0.00000	1.8257	0.23252	0.10910	118	0.23	004.00	0
004	Fiber (%)	0298	1.9000	0.20000	1.8257	0.23252	0.10910	118	0.32	004.00	0
004	Fiber (%)	0425	1.9000	0.00000	1.8257	0.23252	0.10910	118	0.32	004.00	0
004	Fiber (%)	0511	1.9000	0.20000	1.8257	0.23252	0.10910	118	0.32	004.00	0
004	Fiber (%)	0682	1.9000	0.00000	1.8257	0.23252	0.10910	118	0.32	004.07	0
004	Fiber (%)	0354	1.9050	0.13000	1.8257	0.23252	0.10910	118	0.34	004.07	0
004	Fiber (%)	2022	1.9100	0.08000	1.8257	0.23252	0.10910	118	0.36	004.06	0
004	Fiber (%)	2004	1.9200	0.00000	1.8257	0.23252	0.10910	118	0.41	004.00	0
004	Fiber (%)	2149	1.9250	0.03000	1.8257	0.23252	0.10910	118	0.43	004.06	0
004	Fiber (%)	0300	1.9250	0.05000	1.8257	0.23252	0.10910	118	0.43	004.07	0
004	Fiber (%)	0553	1.9250	0.19000	1.8257	0.23252	0.10910	118	0.43	004.07	0
004	Fiber (%)	0205	1.9300	0.08000	1.8257	0.23252	0.10910	118	0.45	004.06	0
004	Fiber (%)	0895	1.9300	0.04000	1.8257	0.23252	0.10910	118	0.45	004.06	0
004	Fiber (%)	0345	1.9300	0.14000	1.8257	0.23252	0.10910	118	0.45	004.00	0
004	Fiber (%)	2093	1.9350	0.01000	1.8257	0.23252	0.10910	118	0.47	004.06	0
004	Fiber (%)	0028	1.9500	0.10000	1.8257	0.23252	0.10910	118	0.53	004.07	0
004	Fiber (%)	0520	1.9500	0.10000	1.8257	0.23252	0.10910	118	0.53	004.07	0
004	Fiber (%)	0964	1.9550	0.17000	1.8257	0.23252	0.10910	118	0.56	004.03	0
004	Fiber (%)	2109	1.9550	0.09000	1.8257	0.23252	0.10910	118	0.56	004.07	0
004	Fiber (%)	0903	1.9750	0.01000	1.8257	0.23252	0.10910	118	0.64	004.06	0
004	Fiber (%)	0074	1.9950	0.05000	1.8257	0.23252	0.10910	118	0.73	004.07	0
004	Fiber (%)	0626	2.0100	0.22000	1.8257	0.23252	0.10910	118	0.79	004.03	0
004	Fiber (%)	0175	2.0200	0.12000	1.8257	0.23252	0.10910	118	0.84	004.00	0
004	Fiber (%)	0004	2.0200	0.22000	1.8257	0.23252	0.10910	118	0.84	004.07	0
004	Fiber (%)	0563	2.0266	0.06940	1.8257	0.23252	0.10910	118	0.86	004.07	0
004	Fiber (%)	0918	2.0350	0.09000	1.8257	0.23252	0.10910	118	0.90	004.06	0
004	Fiber (%)	0045	2.0500	0.18000	1.8257	0.23252	0.10910	118	0.96	004.03	0
004	Fiber (%)	0955	2.0500	0.14000	1.8257	0.23252	0.10910	118	0.96	004.07	0
004	Fiber (%)	0675	2.0900	0.32000	1.8257	0.23252	0.10910	118	1.14	004.06	0
004	Fiber (%)	0226	2.1000	0.40000	1.8257	0.23252	0.10910	118	1.18	004.00	0
004	Fiber (%)	2076	2.1200	0.05000	1.8257	0.23252	0.10910	118	1.27	004.00	0
004	Fiber (%)	0644	2.1305	0.29900	1.8257	0.23252	0.10910	118	1.31	004.07	0
004	Fiber (%)	0536	2.1950	0.05000	1.8257	0.23252	0.10910	118	1.59	004.07	0
004	Fiber (%)	0208	2.2000	0.10000	1.8257	0.23252	0.10910	118	1.61	004.00	0
004	Fiber (%)	0848	2.2150	0.29000	1.8257	0.23252	0.10910	118	1.67	004.07	0
004	Fiber (%)	2062	2.2192	0.03160	1.8257	0.23252	0.10910	118	1.69	004.06	0
004	Fiber (%)	2006	2.2400	0.18000	1.8257	0.23252	0.10910	118	1.78	004.06	0
004	Fiber (%)	0876	2.3050	0.25000	1.8257	0.23252	0.10910	118	2.06	004.07	0
004	Fiber (%)	0878	2.3300	0.14000	1.8257	0.23252	0.10910	118	2.17	004.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber (%)	0089	2.5200	0.00000	1.8257	0.23252	0.10910	118	2.99	004.07	0
004	Fiber (%)	0728	2.8000	0.00000	1.8257	0.23252	0.10910	118	4.19	004.06	0
004	Fiber (%)	0505	2.8300	0.06000	1.8257	0.23252	0.10910	118	4.32	004.07	0
004	Fiber (%)	0676	2.8600	0.18000	1.8257	0.23252	0.10910	118	4.45	004.06	0
004	Fiber (%)	0407	3.2656	0.10900	1.8257	0.23252	0.10910	118	6.19	004.07	0
004	Fiber (%)	0952	6.2150	0.01000	1.8257	0.23252	0.10910	118	18.88	004.06	0
004	Fiber (%)	0366	2.9000	0.80000	1.8257	0.23252	0.10910	118	4.62	004.01	1
004	Fiber (%)	0889	9.9200	0.24000	1.8257	0.23252	0.10910	118	34.81	004.11	2
005	Ash (%)	0683	0.20000	0.02000	8.0128	0.14377	0.05981	138	-54.34	005.04	0
005	Ash (%)	0026	7.6500	0.10000	8.0128	0.14377	0.05981	138	-2.52	005.00	0
005	Ash (%)	0278	7.6600	0.30000	8.0128	0.14377	0.05981	138	-2.45	005.00	0
005	Ash (%)	0226	7.7000	0.20000	8.0128	0.14377	0.05981	138	-2.18	005.00	0
005	Ash (%)	0169	7.7550	0.03000	8.0128	0.14377	0.05981	138	-1.79	005.00	0
005	Ash (%)	0049	7.7650	0.09000	8.0128	0.14377	0.05981	138	-1.72	005.00	0
005	Ash (%)	0004	7.7700	0.04000	8.0128	0.14377	0.05981	138	-1.69	005.00	0
005	Ash (%)	0539	7.7900	0.14000	8.0128	0.14377	0.05981	138	-1.55	005.00	0
005	Ash (%)	0552	7.8150	0.21000	8.0128	0.14377	0.05981	138	-1.38	005.00	0
005	Ash (%)	0623	7.8175	0.01300	8.0128	0.14377	0.05981	138	-1.36	005.00	0
005	Ash (%)	2097	7.8260	0.02080	8.0128	0.14377	0.05981	138	-1.30	005.00	0
005	Ash (%)	0027	7.8295	0.03900	8.0128	0.14377	0.05981	138	-1.28	005.00	0
005	Ash (%)	0019	7.8350	0.07000	8.0128	0.14377	0.05981	138	-1.24	005.00	0
005	Ash (%)	0358	7.8350	0.01000	8.0128	0.14377	0.05981	138	-1.24	005.00	0
005	Ash (%)	0529	7.8400	0.10000	8.0128	0.14377	0.05981	138	-1.20	005.00	0
005	Ash (%)	0861	7.8400	0.04000	8.0128	0.14377	0.05981	138	-1.20	005.00	0
005	Ash (%)	0208	7.8450	0.01000	8.0128	0.14377	0.05981	138	-1.17	005.00	0
005	Ash (%)	0175	7.8550	0.03000	8.0128	0.14377	0.05981	138	-1.10	005.00	0
005	Ash (%)	0511	7.8550	0.01000	8.0128	0.14377	0.05981	138	-1.10	005.00	0
005	Ash (%)	2073	7.8650	0.03000	8.0128	0.14377	0.05981	138	-1.03	005.00	0
005	Ash (%)	0726	7.8650	0.03000	8.0128	0.14377	0.05981	138	-1.03	005.05	0
005	Ash (%)	0693	7.8677	0.04520	8.0128	0.14377	0.05981	138	-1.01	005.00	0
005	Ash (%)	0042	7.8700	0.06000	8.0128	0.14377	0.05981	138	-0.99	005.00	0
005	Ash (%)	0001	7.8746	0.04280	8.0128	0.14377	0.05981	138	-0.96	005.00	0
005	Ash (%)	0298	7.8800	0.02000	8.0128	0.14377	0.05981	138	-0.92	005.00	0
005	Ash (%)	0407	7.8843	0.00010	8.0128	0.14377	0.05981	138	-0.89	005.00	0
005	Ash (%)	0066	7.8850	0.05000	8.0128	0.14377	0.05981	138	-0.89	005.00	0
005	Ash (%)	0144	7.8850	0.07000	8.0128	0.14377	0.05981	138	-0.89	005.00	0
005	Ash (%)	0186	7.8850	0.03000	8.0128	0.14377	0.05981	138	-0.89	005.00	0
005	Ash (%)	0870	7.8939	0.07050	8.0128	0.14377	0.05981	138	-0.83	005.00	0
005	Ash (%)	0148	7.9000	0.06000	8.0128	0.14377	0.05981	138	-0.78	005.00	0
005	Ash (%)	0510	7.9000	0.02000	8.0128	0.14377	0.05981	138	-0.78	005.00	0
005	Ash (%)	0956	7.9000	0.00000	8.0128	0.14377	0.05981	138	-0.78	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0574	7.9050	0.05000	8.0128	0.14377	0.05981	138	-0.75	005.99	0
005	Ash (%)	0968	7.9100	0.04200	8.0128	0.14377	0.05981	138	-0.72	005.00	0
005	Ash (%)	0541	7.9200	0.08000	8.0128	0.14377	0.05981	138	-0.65	005.00	0
005	Ash (%)	0682	7.9200	0.00000	8.0128	0.14377	0.05981	138	-0.65	005.00	0
005	Ash (%)	0194	7.9300	0.02000	8.0128	0.14377	0.05981	138	-0.58	005.00	0
005	Ash (%)	0881	7.9345	0.04900	8.0128	0.14377	0.05981	138	-0.54	005.00	0
005	Ash (%)	0108	7.9350	0.01000	8.0128	0.14377	0.05981	138	-0.54	005.00	0
005	Ash (%)	0553	7.9400	0.06000	8.0128	0.14377	0.05981	138	-0.51	005.00	0
005	Ash (%)	0650	7.9400	0.02000	8.0128	0.14377	0.05981	138	-0.51	005.00	0
005	Ash (%)	0918	7.9400	0.06000	8.0128	0.14377	0.05981	138	-0.51	005.00	0
005	Ash (%)	0205	7.9410	0.02600	8.0128	0.14377	0.05981	138	-0.50	005.00	0
005	Ash (%)	0083	7.9450	0.03000	8.0128	0.14377	0.05981	138	-0.47	005.00	0
005	Ash (%)	0849	7.9495	0.01900	8.0128	0.14377	0.05981	138	-0.44	005.05	0
005	Ash (%)	1019	7.9500	0.14000	8.0128	0.14377	0.05981	138	-0.44	005.00	0
005	Ash (%)	0589	7.9500	0.22000	8.0128	0.14377	0.05981	138	-0.44	005.00	0
005	Ash (%)	0878	7.9500	0.08000	8.0128	0.14377	0.05981	138	-0.44	005.05	0
005	Ash (%)	0265	7.9600	0.02000	8.0128	0.14377	0.05981	138	-0.37	005.00	0
005	Ash (%)	0035	7.9700	0.02000	8.0128	0.14377	0.05981	138	-0.30	005.00	0
005	Ash (%)	0003	7.9700	0.02000	8.0128	0.14377	0.05981	138	-0.30	005.05	0
005	Ash (%)	0598	7.9750	0.01000	8.0128	0.14377	0.05981	138	-0.26	005.00	0
005	Ash (%)	0546	7.9750	0.03000	8.0128	0.14377	0.05981	138	-0.26	005.99	0
005	Ash (%)	0618	7.9755	0.06810	8.0128	0.14377	0.05981	138	-0.26	005.00	0
005	Ash (%)	0622	7.9780	0.02080	8.0128	0.14377	0.05981	138	-0.24	005.00	0
005	Ash (%)	0142	7.9800	0.00000	8.0128	0.14377	0.05981	138	-0.23	005.00	0
005	Ash (%)	0187	7.9800	0.00000	8.0128	0.14377	0.05981	138	-0.23	005.00	0
005	Ash (%)	0425	7.9800	0.02000	8.0128	0.14377	0.05981	138	-0.23	005.00	0
005	Ash (%)	0520	7.9800	0.20000	8.0128	0.14377	0.05981	138	-0.23	005.00	0
005	Ash (%)	2006	7.9850	0.07000	8.0128	0.14377	0.05981	138	-0.19	005.00	0
005	Ash (%)	0033	7.9850	0.01000	8.0128	0.14377	0.05981	138	-0.19	005.00	0
005	Ash (%)	0885	7.9950	0.05000	8.0128	0.14377	0.05981	138	-0.12	005.05	0
005	Ash (%)	2093	7.9950	0.01000	8.0128	0.14377	0.05981	138	-0.12	005.05	0
005	Ash (%)	0630	8.0000	0.04000	8.0128	0.14377	0.05981	138	-0.09	005.00	0
005	Ash (%)	0309	8.0020	0.02800	8.0128	0.14377	0.05981	138	-0.08	005.00	0
005	Ash (%)	2053	8.0050	0.03000	8.0128	0.14377	0.05981	138	-0.05	005.00	0
005	Ash (%)	0692	8.0050	0.13000	8.0128	0.14377	0.05981	138	-0.05	005.05	0
005	Ash (%)	0848	8.0100	0.02000	8.0128	0.14377	0.05981	138	-0.02	005.00	0
005	Ash (%)	2068	8.0100	0.02000	8.0128	0.14377	0.05981	138	-0.02	005.00	0
005	Ash (%)	0563	8.0141	0.01230	8.0128	0.14377	0.05981	138	0.01	005.00	0
005	Ash (%)	0581	8.0150	0.07000	8.0128	0.14377	0.05981	138	0.02	005.00	0
005	Ash (%)	0876	8.0200	0.02000	8.0128	0.14377	0.05981	138	0.05	005.00	0
005	Ash (%)	2150	8.0250	0.05000	8.0128	0.14377	0.05981	138	0.08	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0263	8.0299	0.00660	8.0128	0.14377	0.05981	138	0.12	005.05	0
005	Ash (%)	0940	8.0300	0.06000	8.0128	0.14377	0.05981	138	0.12	005.05	0
005	Ash (%)	0651	8.0325	0.00900	8.0128	0.14377	0.05981	138	0.14	005.00	0
005	Ash (%)	0953	8.0350	0.17000	8.0128	0.14377	0.05981	138	0.15	005.00	0
005	Ash (%)	2022	8.0400	0.06000	8.0128	0.14377	0.05981	138	0.19	005.00	0
005	Ash (%)	0728	8.0400	0.18000	8.0128	0.14377	0.05981	138	0.19	005.99	0
005	Ash (%)	0504	8.0450	0.07000	8.0128	0.14377	0.05981	138	0.22	005.00	0
005	Ash (%)	0047	8.0500	0.10000	8.0128	0.14377	0.05981	138	0.26	005.00	0
005	Ash (%)	0152	8.0500	0.10000	8.0128	0.14377	0.05981	138	0.26	005.00	0
005	Ash (%)	0366	8.0500	0.10000	8.0128	0.14377	0.05981	138	0.26	005.00	0
005	Ash (%)	0893	8.0500	0.00000	8.0128	0.14377	0.05981	138	0.26	005.05	0
005	Ash (%)	0139	8.0550	0.21000	8.0128	0.14377	0.05981	138	0.29	005.00	0
005	Ash (%)	2012	8.0550	0.09000	8.0128	0.14377	0.05981	138	0.29	005.00	0
005	Ash (%)	0683	8.0550	0.01000	8.0128	0.14377	0.05981	138	0.29	005.05	0
005	Ash (%)	2109	8.0550	0.01000	8.0128	0.14377	0.05981	138	0.29	005.05	0
005	Ash (%)	0653	8.0650	0.01000	8.0128	0.14377	0.05981	138	0.36	005.00	0
005	Ash (%)	0164	8.0700	0.04000	8.0128	0.14377	0.05981	138	0.40	005.00	0
005	Ash (%)	0353	8.0700	0.22000	8.0128	0.14377	0.05981	138	0.40	005.00	0
005	Ash (%)	0620	8.0705	0.00370	8.0128	0.14377	0.05981	138	0.40	005.00	0
005	Ash (%)	0098	8.0750	0.05000	8.0128	0.14377	0.05981	138	0.43	005.00	0
005	Ash (%)	0505	8.0750	0.05000	8.0128	0.14377	0.05981	138	0.43	005.00	0
005	Ash (%)	0202	8.0750	0.01000	8.0128	0.14377	0.05981	138	0.43	005.99	0
005	Ash (%)	0939	8.0800	0.04000	8.0128	0.14377	0.05981	138	0.47	005.05	0
005	Ash (%)	0619	8.0850	0.01000	8.0128	0.14377	0.05981	138	0.50	005.05	0
005	Ash (%)	2062	8.0879	0.02050	8.0128	0.14377	0.05981	138	0.52	005.05	0
005	Ash (%)	0590	8.0900	0.00000	8.0128	0.14377	0.05981	138	0.54	005.05	0
005	Ash (%)	0354	8.0950	0.03000	8.0128	0.14377	0.05981	138	0.57	005.00	0
005	Ash (%)	0062	8.1000	0.00000	8.0128	0.14377	0.05981	138	0.61	005.00	0
005	Ash (%)	0294	8.1000	0.00000	8.0128	0.14377	0.05981	138	0.61	005.05	0
005	Ash (%)	0413	8.1000	0.20000	8.0128	0.14377	0.05981	138	0.61	005.05	0
005	Ash (%)	2004	8.1000	0.08000	8.0128	0.14377	0.05981	138	0.61	005.99	0
005	Ash (%)	0345	8.1200	0.04000	8.0128	0.14377	0.05981	138	0.75	005.00	0
005	Ash (%)	0629	8.1200	0.02000	8.0128	0.14377	0.05981	138	0.75	005.99	0
005	Ash (%)	0100	8.1250	0.27000	8.0128	0.14377	0.05981	138	0.78	005.00	0
005	Ash (%)	2145	8.1300	0.00000	8.0128	0.14377	0.05981	138	0.82	005.05	0
005	Ash (%)	0689	8.1350	0.01000	8.0128	0.14377	0.05981	138	0.85	005.05	0
005	Ash (%)	0903	8.1350	0.01000	8.0128	0.14377	0.05981	138	0.85	005.05	0
005	Ash (%)	0615	8.1400	0.10000	8.0128	0.14377	0.05981	138	0.88	005.00	0
005	Ash (%)	0897	8.1450	0.01000	8.0128	0.14377	0.05981	138	0.92	005.05	0
005	Ash (%)	0685	8.1500	0.00000	8.0128	0.14377	0.05981	138	0.95	005.05	0
005	Ash (%)	0910	8.1500	0.30000	8.0128	0.14377	0.05981	138	0.95	005.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0889	8.1550	0.01000	8.0128	0.14377	0.05981	138	0.99	005.05	0
005	Ash (%)	0955	8.1550	0.03000	8.0128	0.14377	0.05981	138	0.99	005.05	0
005	Ash (%)	2152	8.1600	0.02000	8.0128	0.14377	0.05981	138	1.02	005.05	0
005	Ash (%)	0610	8.1650	0.01000	8.0128	0.14377	0.05981	138	1.06	005.05	0
005	Ash (%)	2089	8.1800	0.00000	8.0128	0.14377	0.05981	138	1.16	005.00	0
005	Ash (%)	0895	8.1800	0.02000	8.0128	0.14377	0.05981	138	1.16	005.05	0
005	Ash (%)	0921	8.1800	0.00000	8.0128	0.14377	0.05981	138	1.16	005.05	0
005	Ash (%)	2009	8.1832	0.02080	8.0128	0.14377	0.05981	138	1.19	005.05	0
005	Ash (%)	0652	8.2000	0.00000	8.0128	0.14377	0.05981	138	1.30	005.99	0
005	Ash (%)	0948	8.2050	0.01000	8.0128	0.14377	0.05981	138	1.34	005.00	0
005	Ash (%)	0190	8.2300	0.04000	8.0128	0.14377	0.05981	138	1.51	005.05	0
005	Ash (%)	2129	8.2300	0.12000	8.0128	0.14377	0.05981	138	1.51	005.05	0
005	Ash (%)	0171	8.2400	0.04000	8.0128	0.14377	0.05981	138	1.58	005.00	0
005	Ash (%)	2149	8.2410	0.00600	8.0128	0.14377	0.05981	138	1.59	005.05	0
005	Ash (%)	0045	8.2450	0.27000	8.0128	0.14377	0.05981	138	1.61	005.00	0
005	Ash (%)	2146	8.2750	0.01000	8.0128	0.14377	0.05981	138	1.82	005.05	0
005	Ash (%)	0675	8.2900	0.30000	8.0128	0.14377	0.05981	138	1.93	005.00	0
005	Ash (%)	0644	8.2900	0.10000	8.0128	0.14377	0.05981	138	1.93	005.02	0
005	Ash (%)	0357	8.3000	0.00000	8.0128	0.14377	0.05981	138	2.00	005.00	0
005	Ash (%)	0941	8.3050	0.11000	8.0128	0.14377	0.05981	138	2.03	005.05	0
005	Ash (%)	0676	8.4100	0.00000	8.0128	0.14377	0.05981	138	2.76	005.99	0
005	Ash (%)	0297	8.8100	0.46000	8.0128	0.14377	0.05981	138	5.54	005.11	0
005	Ash (%)	0874	9.8500	0.10000	8.0128	0.14377	0.05981	138	12.78	005.00	0
005	Ash (%)	0889	6.3550	0.99000	8.0128	0.14377	0.05981	138	-11.53	005.11	1
005	Ash (%)	0297	7.5500	0.90000	8.0128	0.14377	0.05981	138	-3.22	005.00	1
005	Ash (%)	2075	8.3450	0.63000	8.0128	0.14377	0.05981	138	2.31	005.02	1
005	Ash (%)	0536	25.440	0.20000	8.0128	0.14377	0.05981	138	121.22	005.99	2
006	Total sugars (%)	0910	19.150	0.68000	20.366	2.2177	0.39825	4	-0.58	006.00	0
006	Total sugars (%)	0956	19.150	0.70000	20.366	2.2177	0.39825	4	-0.58	006.99	0
006	Total sugars (%)	0921	19.480	0.00000	20.366	2.2177	0.39825	4	1.15	006.00	0
006	Total sugars (%)	0407	23.685	0.21300	20.366	2.2177	0.39825	4	23.22	006.01	0
008	Fiber (%)	2012	1.6300	0.06000	2.6804	0.45617	0.20951	60	-2.30	008.08	0
008	Fiber (%)	0676	1.9300	0.04000	2.6804	0.45617	0.20951	60	-1.64	008.02	0
008	Fiber (%)	0354	1.9550	0.15000	2.6804	0.45617	0.20951	60	-1.59	008.08	0
008	Fiber (%)	0918	1.9765	0.74100	2.6804	0.45617	0.20951	60	-1.54	008.08	0
008	Fiber (%)	0035	2.1950	0.07000	2.6804	0.45617	0.20951	60	-1.06	008.08	0
008	Fiber (%)	0590	2.2000	0.00000	2.6804	0.45617	0.20951	60	-1.05	008.02	0
008	Fiber (%)	0910	2.2450	0.09000	2.6804	0.45617	0.20951	60	-0.95	008.08	0
008	Fiber (%)	0004	2.2700	0.18000	2.6804	0.45617	0.20951	60	-0.90	008.08	0
008	Fiber (%)	2004	2.3050	0.07000	2.6804	0.45617	0.20951	60	-0.82	008.08	0
008	Fiber (%)	0619	2.3120	0.22800	2.6804	0.45617	0.20951	60	-0.81	008.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
008	Fiber (%)	0889	2.3250	0.61000	2.6804	0.45617	0.20951	60	-0.78	008.08	0
008	Fiber (%)	0653	2.3450	0.05000	2.6804	0.45617	0.20951	60	-0.74	008.08	0
008	Fiber (%)	0042	2.3800	0.22000	2.6804	0.45617	0.20951	60	-0.66	008.08	0
008	Fiber (%)	0098	2.3900	0.36000	2.6804	0.45617	0.20951	60	-0.64	008.02	0
008	Fiber (%)	0861	2.3950	0.09000	2.6804	0.45617	0.20951	60	-0.63	008.08	0
008	Fiber (%)	0413	2.4000	0.20000	2.6804	0.45617	0.20951	60	-0.61	008.08	0
008	Fiber (%)	0511	2.4000	0.20000	2.6804	0.45617	0.20951	60	-0.61	008.99	0
008	Fiber (%)	0948	2.4100	0.00000	2.6804	0.45617	0.20951	60	-0.59	008.02	0
008	Fiber (%)	0358	2.4300	0.18000	2.6804	0.45617	0.20951	60	-0.55	008.08	0
008	Fiber (%)	0049	2.4350	0.25000	2.6804	0.45617	0.20951	60	-0.54	008.08	0
008	Fiber (%)	0190	2.4400	0.08000	2.6804	0.45617	0.20951	60	-0.53	008.08	0
008	Fiber (%)	0878	2.4400	0.06000	2.6804	0.45617	0.20951	60	-0.53	008.08	0
008	Fiber (%)	0581	2.4450	0.23000	2.6804	0.45617	0.20951	60	-0.52	008.08	0
008	Fiber (%)	0164	2.4500	0.10000	2.6804	0.45617	0.20951	60	-0.51	008.08	0
008	Fiber (%)	0357	2.4500	0.30000	2.6804	0.45617	0.20951	60	-0.51	008.08	0
008	Fiber (%)	0001	2.4800	0.38000	2.6804	0.45617	0.20951	60	-0.44	008.08	0
008	Fiber (%)	0345	2.5250	0.15000	2.6804	0.45617	0.20951	60	-0.34	008.02	0
008	Fiber (%)	0294	2.5500	0.10000	2.6804	0.45617	0.20951	60	-0.29	008.08	0
008	Fiber (%)	0037	2.6150	0.71000	2.6804	0.45617	0.20951	60	-0.14	008.08	0
008	Fiber (%)	0309	2.6215	0.03550	2.6804	0.45617	0.20951	60	-0.13	008.02	0
008	Fiber (%)	0693	2.6269	0.22270	2.6804	0.45617	0.20951	60	-0.12	008.08	0
008	Fiber (%)	0504	2.6700	0.36000	2.6804	0.45617	0.20951	60	-0.02	008.02	0
008	Fiber (%)	0529	2.6750	0.27000	2.6804	0.45617	0.20951	60	-0.01	008.08	0
008	Fiber (%)	0726	2.6850	0.07000	2.6804	0.45617	0.20951	60	0.01	008.08	0
008	Fiber (%)	0148	2.7100	0.08000	2.6804	0.45617	0.20951	60	0.06	008.08	0
008	Fiber (%)	0208	2.7300	0.40000	2.6804	0.45617	0.20951	60	0.11	008.02	0
008	Fiber (%)	0226	2.7500	0.50000	2.6804	0.45617	0.20951	60	0.15	008.02	0
008	Fiber (%)	0148	2.7700	0.02000	2.6804	0.45617	0.20951	60	0.20	008.02	0
008	Fiber (%)	0675	2.7800	0.04000	2.6804	0.45617	0.20951	60	0.22	008.02	0
008	Fiber (%)	0870	2.8095	0.10740	2.6804	0.45617	0.20951	60	0.28	008.08	0
008	Fiber (%)	0083	2.8450	0.21000	2.6804	0.45617	0.20951	60	0.36	008.08	0
008	Fiber (%)	0939	2.8700	0.18000	2.6804	0.45617	0.20951	60	0.42	008.02	0
008	Fiber (%)	0297	2.8700	0.10000	2.6804	0.45617	0.20951	60	0.42	008.08	0
008	Fiber (%)	0026	2.9550	0.11000	2.6804	0.45617	0.20951	60	0.60	008.08	0
008	Fiber (%)	0265	3.0000	0.40000	2.6804	0.45617	0.20951	60	0.70	008.05	0
008	Fiber (%)	0956	3.0000	0.00000	2.6804	0.45617	0.20951	60	0.70	008.08	0
008	Fiber (%)	0968	3.0430	0.15200	2.6804	0.45617	0.20951	60	0.79	008.08	0
008	Fiber (%)	0263	3.0649	0.23660	2.6804	0.45617	0.20951	60	0.84	008.08	0
008	Fiber (%)	0045	3.0650	0.13000	2.6804	0.45617	0.20951	60	0.84	008.08	0
008	Fiber (%)	0353	3.0850	0.31000	2.6804	0.45617	0.20951	60	0.89	008.02	0
008	Fiber (%)	0278	3.2500	0.10000	2.6804	0.45617	0.20951	60	1.25	008.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
008	Fiber (%)	0407	3.2787	0.00720	2.6804	0.45617	0.20951	60	1.31	008.08	0
008	Fiber (%)	0366	3.3000	0.40000	2.6804	0.45617	0.20951	60	1.36	008.08	0
008	Fiber (%)	0510	3.4500	0.10000	2.6804	0.45617	0.20951	60	1.69	008.08	0
008	Fiber (%)	0848	3.5000	0.54000	2.6804	0.45617	0.20951	60	1.80	008.08	0
008	Fiber (%)	2129	4.0600	0.36000	2.6804	0.45617	0.20951	60	3.02	008.08	0
008	Fiber (%)	0187	4.0700	0.02000	2.6804	0.45617	0.20951	60	3.05	008.02	0
008	Fiber (%)	0610	4.0750	0.55000	2.6804	0.45617	0.20951	60	3.06	008.99	0
008	Fiber (%)	0169	4.3100	0.52000	2.6804	0.45617	0.20951	60	3.57	008.08	0
008	Fiber (%)	0728	5.3300	0.14000	2.6804	0.45617	0.20951	60	5.81	008.02	0
008	Fiber (%)	0536	9.4500	0.04000	2.6804	0.45617	0.20951	60	14.84	008.08	2
009	Fiber (%)	0889	4.6600	0.40000	6.1358	1.1497	0.29225	57	-1.28	009.09	0
009	Fiber (%)	2004	4.6900	0.00000	6.1358	1.1497	0.29225	57	-1.26	009.09	0
009	Fiber (%)	0164	4.8000	0.20000	6.1358	1.1497	0.29225	57	-1.16	009.09	0
009	Fiber (%)	0354	4.8650	0.03000	6.1358	1.1497	0.29225	57	-1.11	009.09	0
009	Fiber (%)	0598	5.0150	0.11000	6.1358	1.1497	0.29225	57	-0.97	009.09	0
009	Fiber (%)	0357	5.1000	0.00000	6.1358	1.1497	0.29225	57	-0.90	009.09	0
009	Fiber (%)	0413	5.1000	0.20000	6.1358	1.1497	0.29225	57	-0.90	009.09	0
009	Fiber (%)	0098	5.1050	0.13000	6.1358	1.1497	0.29225	57	-0.90	009.07	0
009	Fiber (%)	2129	5.1050	0.37000	6.1358	1.1497	0.29225	57	-0.90	009.09	0
009	Fiber (%)	0187	5.1200	0.16000	6.1358	1.1497	0.29225	57	-0.88	009.07	0
009	Fiber (%)	0358	5.2450	0.13000	6.1358	1.1497	0.29225	57	-0.77	009.09	0
009	Fiber (%)	0294	5.3000	0.00000	6.1358	1.1497	0.29225	57	-0.73	009.09	0
009	Fiber (%)	0148	5.3350	0.09000	6.1358	1.1497	0.29225	57	-0.70	009.09	0
009	Fiber (%)	0861	5.3650	0.01000	6.1358	1.1497	0.29225	57	-0.67	009.09	0
009	Fiber (%)	0870	5.4005	0.11360	6.1358	1.1497	0.29225	57	-0.64	009.09	0
009	Fiber (%)	0693	5.4107	0.07250	6.1358	1.1497	0.29225	57	-0.63	009.09	0
009	Fiber (%)	0878	5.4450	0.11000	6.1358	1.1497	0.29225	57	-0.60	009.09	0
009	Fiber (%)	0918	5.5215	0.05700	6.1358	1.1497	0.29225	57	-0.53	009.09	0
009	Fiber (%)	0619	5.5305	0.58100	6.1358	1.1497	0.29225	57	-0.53	009.09	0
009	Fiber (%)	0353	5.5550	0.35000	6.1358	1.1497	0.29225	57	-0.51	009.07	0
009	Fiber (%)	0263	5.5900	0.08670	6.1358	1.1497	0.29225	57	-0.47	009.09	0
009	Fiber (%)	0581	5.6250	0.31000	6.1358	1.1497	0.29225	57	-0.44	009.09	0
009	Fiber (%)	0939	5.6900	0.24000	6.1358	1.1497	0.29225	57	-0.39	009.09	0
009	Fiber (%)	0407	5.6950	0.03600	6.1358	1.1497	0.29225	57	-0.38	009.09	0
009	Fiber (%)	0956	5.7000	0.20000	6.1358	1.1497	0.29225	57	-0.38	009.09	0
009	Fiber (%)	0037	5.7300	0.02000	6.1358	1.1497	0.29225	57	-0.35	009.09	0
009	Fiber (%)	0049	5.7300	0.26000	6.1358	1.1497	0.29225	57	-0.35	009.09	0
009	Fiber (%)	0504	5.7350	0.01000	6.1358	1.1497	0.29225	57	-0.35	009.04	0
009	Fiber (%)	0001	5.7500	0.30000	6.1358	1.1497	0.29225	57	-0.34	009.09	0
009	Fiber (%)	0045	5.7800	0.08000	6.1358	1.1497	0.29225	57	-0.31	009.09	0
009	Fiber (%)	0083	5.8150	0.17000	6.1358	1.1497	0.29225	57	-0.28	009.09	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
009	Fiber (%)	0297	6.0400	0.16000	6.1358	1.1497	0.29225	57	-0.08	009.09	0
009	Fiber (%)	0345	6.0500	0.30000	6.1358	1.1497	0.29225	57	-0.07	009.07	0
009	Fiber (%)	2022	6.0650	0.01000	6.1358	1.1497	0.29225	57	-0.06	009.09	0
009	Fiber (%)	0265	6.1000	0.40000	6.1358	1.1497	0.29225	57	-0.03	009.09	0
009	Fiber (%)	0278	6.1500	0.70000	6.1358	1.1497	0.29225	57	0.01	009.09	0
009	Fiber (%)	2012	6.2650	0.25000	6.1358	1.1497	0.29225	57	0.11	009.09	0
009	Fiber (%)	0366	6.3000	0.20000	6.1358	1.1497	0.29225	57	0.14	009.09	0
009	Fiber (%)	0948	6.3250	0.59000	6.1358	1.1497	0.29225	57	0.16	009.09	0
009	Fiber (%)	0675	6.3950	0.23000	6.1358	1.1497	0.29225	57	0.23	009.07	0
009	Fiber (%)	0910	6.4500	0.10000	6.1358	1.1497	0.29225	57	0.27	009.07	0
009	Fiber (%)	0510	6.5000	0.80000	6.1358	1.1497	0.29225	57	0.32	009.09	0
009	Fiber (%)	0169	6.9250	0.13000	6.1358	1.1497	0.29225	57	0.69	009.09	0
009	Fiber (%)	0728	7.0450	0.55000	6.1358	1.1497	0.29225	57	0.79	009.99	0
009	Fiber (%)	0676	7.1650	1.0300	6.1358	1.1497	0.29225	57	0.90	009.07	0
009	Fiber (%)	0226	7.3500	0.10000	6.1358	1.1497	0.29225	57	1.06	009.07	0
009	Fiber (%)	0590	7.5500	0.50000	6.1358	1.1497	0.29225	57	1.23	009.07	0
009	Fiber (%)	0726	7.8150	0.15000	6.1358	1.1497	0.29225	57	1.46	009.09	0
009	Fiber (%)	0309	8.0947	0.17070	6.1358	1.1497	0.29225	57	1.70	009.07	0
009	Fiber (%)	0190	8.3700	0.18000	6.1358	1.1497	0.29225	57	1.94	009.09	0
009	Fiber (%)	0208	8.7600	1.2800	6.1358	1.1497	0.29225	57	2.28	009.07	0
009	Fiber (%)	0610	8.9500	1.3000	6.1358	1.1497	0.29225	57	2.45	009.07	0
009	Fiber (%)	0653	9.1700	0.02000	6.1358	1.1497	0.29225	57	2.64	009.09	0
009	Fiber (%)	0848	9.7700	0.52000	6.1358	1.1497	0.29225	57	3.16	009.09	0
009	Fiber (%)	0536	10.030	1.2400	6.1358	1.1497	0.29225	57	3.39	009.09	0
009	Fiber (%)	0889	10.210	0.82000	6.1358	1.1497	0.29225	57	3.54	009.99	0
009	Fiber (%)	0968	11.899	0.10100	6.1358	1.1497	0.29225	57	5.01	009.09	0
010	Moisture (%)	0546	7.8100	0.04000	9.1716	0.67369	0.11842	19	-2.02	010.99	0
010	Moisture (%)	0952	8.5000	0.02000	9.1716	0.67369	0.11842	19	-1.00	010.99	0
010	Moisture (%)	0843	8.6250	0.13000	9.1716	0.67369	0.11842	19	-0.81	010.03	0
010	Moisture (%)	0917	8.6500	0.10000	9.1716	0.67369	0.11842	19	-0.77	010.99	0
010	Moisture (%)	0948	8.6500	0.10000	9.1716	0.67369	0.11842	19	-0.77	010.99	0
010	Moisture (%)	2089	8.6850	0.05000	9.1716	0.67369	0.11842	19	-0.72	010.99	0
010	Moisture (%)	0164	8.8100	0.06000	9.1716	0.67369	0.11842	19	-0.54	010.03	0
010	Moisture (%)	2150	8.8300	0.40000	9.1716	0.67369	0.11842	19	-0.51	010.99	0
010	Moisture (%)	0208	8.9400	0.42000	9.1716	0.67369	0.11842	19	-0.34	010.03	0
010	Moisture (%)	0939	8.9800	0.04000	9.1716	0.67369	0.11842	19	-0.28	010.99	0
010	Moisture (%)	0889	9.2700	0.08000	9.1716	0.67369	0.11842	19	0.15	010.11	0
010	Moisture (%)	0692	9.3500	0.24000	9.1716	0.67369	0.11842	19	0.26	010.99	0
010	Moisture (%)	0405	9.5750	0.01000	9.1716	0.67369	0.11842	19	0.60	010.99	0
010	Moisture (%)	2129	9.6550	0.05000	9.1716	0.67369	0.11842	19	0.72	010.99	0
010	Moisture (%)	0726	9.6650	0.13000	9.1716	0.67369	0.11842	19	0.73	010.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
010	Moisture (%)	2004	9.7000	0.06000	9.1716	0.67369	0.11842	19	0.78	010.99	0
010	Moisture (%)	0190	9.8500	0.08000	9.1716	0.67369	0.11842	19	1.01	010.99	0
010	Moisture (%)	2105	10.430	0.22000	9.1716	0.67369	0.11842	19	1.87	010.99	0
010	Moisture (%)	2127	11.390	0.02000	9.1716	0.67369	0.11842	19	3.29	010.99	0
010	Moisture (%)	0921	9.5950	0.77000	9.1716	0.67369	0.11842	19	0.63	010.99	1
011	Loss on Drying (%)	0874	8.0250	0.07000	11.223	0.70267	0.17200	73	-4.55	011.01	0
011	Loss on Drying (%)	1019	8.7650	0.09000	11.223	0.70267	0.17200	73	-3.50	011.01	0
011	Loss on Drying (%)	0541	9.4900	0.04000	11.223	0.70267	0.17200	73	-2.47	011.99	0
011	Loss on Drying (%)	0345	9.7200	0.04000	11.223	0.70267	0.17200	73	-2.14	011.01	0
011	Loss on Drying (%)	2076	10.070	0.10000	11.223	0.70267	0.17200	73	-1.64	011.99	0
011	Loss on Drying (%)	0574	10.170	0.06000	11.223	0.70267	0.17200	73	-1.50	011.01	0
011	Loss on Drying (%)	0536	10.220	0.22000	11.223	0.70267	0.17200	73	-1.43	011.01	0
011	Loss on Drying (%)	0598	10.330	0.26000	11.223	0.70267	0.17200	73	-1.27	011.01	0
011	Loss on Drying (%)	0942	10.420	0.64000	11.223	0.70267	0.17200	73	-1.14	011.02	0
011	Loss on Drying (%)	0563	10.461	0.08880	11.223	0.70267	0.17200	73	-1.09	011.01	0
011	Loss on Drying (%)	0953	10.465	0.15000	11.223	0.70267	0.17200	73	-1.08	011.01	0
011	Loss on Drying (%)	0175	10.550	0.10000	11.223	0.70267	0.17200	73	-0.96	011.01	0
011	Loss on Drying (%)	0194	10.585	0.03000	11.223	0.70267	0.17200	73	-0.91	011.01	0
011	Loss on Drying (%)	0108	10.605	0.55000	11.223	0.70267	0.17200	73	-0.88	011.01	0
011	Loss on Drying (%)	0859	10.649	0.07300	11.223	0.70267	0.17200	73	-0.82	011.01	0
011	Loss on Drying (%)	0857	10.650	0.10000	11.223	0.70267	0.17200	73	-0.82	011.99	0
011	Loss on Drying (%)	0843	10.720	0.68000	11.223	0.70267	0.17200	73	-0.72	011.02	0
011	Loss on Drying (%)	0622	10.726	0.13030	11.223	0.70267	0.17200	73	-0.71	011.01	0
011	Loss on Drying (%)	0651	10.754	0.09900	11.223	0.70267	0.17200	73	-0.67	011.01	0
011	Loss on Drying (%)	0870	10.804	0.37020	11.223	0.70267	0.17200	73	-0.60	011.01	0
011	Loss on Drying (%)	0653	10.815	0.09000	11.223	0.70267	0.17200	73	-0.58	011.01	0
011	Loss on Drying (%)	0100	10.830	0.10000	11.223	0.70267	0.17200	73	-0.56	011.01	0
011	Loss on Drying (%)	0510	10.875	0.05000	11.223	0.70267	0.17200	73	-0.50	011.01	0
011	Loss on Drying (%)	0226	10.945	0.37000	11.223	0.70267	0.17200	73	-0.40	011.01	0
011	Loss on Drying (%)	2075	10.955	0.09000	11.223	0.70267	0.17200	73	-0.38	011.01	0
011	Loss on Drying (%)	2097	11.023	0.07470	11.223	0.70267	0.17200	73	-0.29	011.01	0
011	Loss on Drying (%)	0675	11.040	0.04000	11.223	0.70267	0.17200	73	-0.26	011.01	0
011	Loss on Drying (%)	0263	11.050	0.10000	11.223	0.70267	0.17200	73	-0.25	011.01	0
011	Loss on Drying (%)	0033	11.100	0.00000	11.223	0.70267	0.17200	73	-0.18	011.01	0
011	Loss on Drying (%)	0425	11.105	0.07000	11.223	0.70267	0.17200	73	-0.17	011.02	0
011	Loss on Drying (%)	0148	11.130	0.16000	11.223	0.70267	0.17200	73	-0.13	011.01	0
011	Loss on Drying (%)	0298	11.160	0.02000	11.223	0.70267	0.17200	73	-0.09	011.01	0
011	Loss on Drying (%)	0294	11.200	0.00000	11.223	0.70267	0.17200	73	-0.03	011.01	0
011	Loss on Drying (%)	0881	11.201	0.17200	11.223	0.70267	0.17200	73	-0.03	011.01	0
011	Loss on Drying (%)	0650	11.215	0.37000	11.223	0.70267	0.17200	73	-0.01	011.01	0
011	Loss on Drying (%)	0682	11.220	0.00000	11.223	0.70267	0.17200	73	0.00	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
011	Loss on Drying (%)	0202	11.230	0.00000	11.223	0.70267	0.17200	73	0.01	011.01	0
011	Loss on Drying (%)	0208	11.250	0.50000	11.223	0.70267	0.17200	73	0.04	011.01	0
011	Loss on Drying (%)	0171	11.265	0.13000	11.223	0.70267	0.17200	73	0.06	011.01	0
011	Loss on Drying (%)	0152	11.300	0.20000	11.223	0.70267	0.17200	73	0.11	011.01	0
011	Loss on Drying (%)	0520	11.300	0.00000	11.223	0.70267	0.17200	73	0.11	011.01	0
011	Loss on Drying (%)	0552	11.340	0.02000	11.223	0.70267	0.17200	73	0.17	011.01	0
011	Loss on Drying (%)	0620	11.375	0.30780	11.223	0.70267	0.17200	73	0.22	011.01	0
011	Loss on Drying (%)	0026	11.375	0.01000	11.223	0.70267	0.17200	73	0.22	011.01	0
011	Loss on Drying (%)	0354	11.375	0.03000	11.223	0.70267	0.17200	73	0.22	011.01	0
011	Loss on Drying (%)	0848	11.385	0.27000	11.223	0.70267	0.17200	73	0.23	011.01	0
011	Loss on Drying (%)	2031	11.418	0.24380	11.223	0.70267	0.17200	73	0.28	011.01	0
011	Loss on Drying (%)	0066	11.430	0.30000	11.223	0.70267	0.17200	73	0.29	011.01	0
011	Loss on Drying (%)	2006	11.465	0.03000	11.223	0.70267	0.17200	73	0.34	011.01	0
011	Loss on Drying (%)	0229	11.510	0.06000	11.223	0.70267	0.17200	73	0.41	011.01	0
011	Loss on Drying (%)	0407	11.511	0.02120	11.223	0.70267	0.17200	73	0.41	011.01	0
011	Loss on Drying (%)	2073	11.560	0.04000	11.223	0.70267	0.17200	73	0.48	011.01	0
011	Loss on Drying (%)	0062	11.600	0.00000	11.223	0.70267	0.17200	73	0.54	011.01	0
011	Loss on Drying (%)	0205	11.625	0.03000	11.223	0.70267	0.17200	73	0.57	011.01	0
011	Loss on Drying (%)	0685	11.625	0.01000	11.223	0.70267	0.17200	73	0.57	011.01	0
011	Loss on Drying (%)	2068	11.685	0.17000	11.223	0.70267	0.17200	73	0.66	011.01	0
011	Loss on Drying (%)	0623	11.688	0.38200	11.223	0.70267	0.17200	73	0.66	011.01	0
011	Loss on Drying (%)	0233	11.710	0.12000	11.223	0.70267	0.17200	73	0.69	011.01	0
011	Loss on Drying (%)	0098	11.715	0.13000	11.223	0.70267	0.17200	73	0.70	011.01	0
011	Loss on Drying (%)	0144	11.785	0.03000	11.223	0.70267	0.17200	73	0.80	011.01	0
011	Loss on Drying (%)	0541	11.785	0.07000	11.223	0.70267	0.17200	73	0.80	011.01	0
011	Loss on Drying (%)	0265	11.850	0.50000	11.223	0.70267	0.17200	73	0.89	011.01	0
011	Loss on Drying (%)	0511	11.940	0.18000	11.223	0.70267	0.17200	73	1.02	011.01	0
011	Loss on Drying (%)	0968	12.062	0.05300	11.223	0.70267	0.17200	73	1.19	011.01	0
011	Loss on Drying (%)	0164	12.115	0.17000	11.223	0.70267	0.17200	73	1.27	011.01	0
011	Loss on Drying (%)	0358	12.305	0.23000	11.223	0.70267	0.17200	73	1.54	011.01	0
011	Loss on Drying (%)	0553	12.410	0.98000	11.223	0.70267	0.17200	73	1.69	011.01	0
011	Loss on Drying (%)	0861	12.420	0.14000	11.223	0.70267	0.17200	73	1.70	011.01	0
011	Loss on Drying (%)	2022	12.520	0.24000	11.223	0.70267	0.17200	73	1.85	011.01	0
011	Loss on Drying (%)	0728	12.560	0.20000	11.223	0.70267	0.17200	73	1.90	011.01	0
011	Loss on Drying (%)	0539	12.730	0.56000	11.223	0.70267	0.17200	73	2.14	011.01	0
011	Loss on Drying (%)	0589	12.945	0.55000	11.223	0.70267	0.17200	73	2.45	011.01	0
011	Loss on Drying (%)	0309	13.435	0.05000	11.223	0.70267	0.17200	73	3.15	011.01	0
011	Loss on Drying (%)	0047	10.400	3.0000	11.223	0.70267	0.17200	73	-1.17	011.01	1
012	Starch (%)	0510	13.000	0.20000	19.231	1.1590	0.24746	30	-5.38	012.04	0
012	Starch (%)	0918	14.296	0.46100	19.231	1.1590	0.24746	30	-4.26	012.01	0
012	Starch (%)	0885	16.446	0.28970	19.231	1.1590	0.24746	30	-2.40	012.03	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
012	Starch (%)	0265	17.150	0.10000	19.231	1.1590	0.24746	30	-1.80	012.01	0
012	Starch (%)	0956	17.900	0.20000	19.231	1.1590	0.24746	30	-1.15	012.01	0
012	Starch (%)	0164	18.050	0.50000	19.231	1.1590	0.24746	30	-1.02	012.00	0
012	Starch (%)	0278	18.550	0.50000	19.231	1.1590	0.24746	30	-0.59	012.04	0
012	Starch (%)	0160	18.705	0.17000	19.231	1.1590	0.24746	30	-0.45	012.04	0
012	Starch (%)	0045	18.750	0.10000	19.231	1.1590	0.24746	30	-0.42	012.01	0
012	Starch (%)	0098	18.900	0.60000	19.231	1.1590	0.24746	30	-0.29	012.03	0
012	Starch (%)	0676	18.930	0.52000	19.231	1.1590	0.24746	30	-0.26	012.01	0
012	Starch (%)	2004	19.000	0.60000	19.231	1.1590	0.24746	30	-0.20	012.01	0
012	Starch (%)	0297	19.010	0.08000	19.231	1.1590	0.24746	30	-0.19	012.03	0
012	Starch (%)	0693	19.136	0.33220	19.231	1.1590	0.24746	30	-0.08	012.01	0
012	Starch (%)	0619	19.150	0.10000	19.231	1.1590	0.24746	30	-0.07	012.00	0
012	Starch (%)	0407	19.363	0.06500	19.231	1.1590	0.24746	30	0.11	012.03	0
012	Starch (%)	0354	19.435	0.23000	19.231	1.1590	0.24746	30	0.18	012.00	0
012	Starch (%)	0870	19.612	0.21770	19.231	1.1590	0.24746	30	0.33	012.01	0
012	Starch (%)	0297	19.640	0.84000	19.231	1.1590	0.24746	30	0.35	012.11	0
012	Starch (%)	0878	19.750	0.10000	19.231	1.1590	0.24746	30	0.45	012.00	0
012	Starch (%)	0164	19.750	0.30000	19.231	1.1590	0.24746	30	0.45	012.01	0
012	Starch (%)	0683	19.920	0.16000	19.231	1.1590	0.24746	30	0.59	012.00	0
012	Starch (%)	0889	20.145	0.07000	19.231	1.1590	0.24746	30	0.79	012.11	0
012	Starch (%)	0689	20.150	0.30000	19.231	1.1590	0.24746	30	0.79	012.00	0
012	Starch (%)	0910	20.150	0.10000	19.231	1.1590	0.24746	30	0.79	012.00	0
012	Starch (%)	2009	20.161	0.04820	19.231	1.1590	0.24746	30	0.80	012.00	0
012	Starch (%)	0889	20.340	0.00000	19.231	1.1590	0.24746	30	0.96	012.03	0
012	Starch (%)	0948	20.535	0.01000	19.231	1.1590	0.24746	30	1.12	012.04	0
012	Starch (%)	2006	24.655	0.01000	19.231	1.1590	0.24746	30	4.68	012.00	0
012	Starch (%)	0552	24.780	0.22000	19.231	1.1590	0.24746	30	4.79	012.00	0
013	Fat (%)	0968	6.5510	0.03800	8.1562	0.79175	0.16437	45	-2.03	013.13	0
013	Fat (%)	0047	6.8800	0.02000	8.1562	0.79175	0.16437	45	-1.61	013.08	0
013	Fat (%)	0083	7.1500	0.20000	8.1562	0.79175	0.16437	45	-1.27	013.00	0
013	Fat (%)	2073	7.1600	0.02000	8.1562	0.79175	0.16437	45	-1.26	013.02	0
013	Fat (%)	0581	7.2300	0.16000	8.1562	0.79175	0.16437	45	-1.17	013.13	0
013	Fat (%)	2145	7.2450	0.03000	8.1562	0.79175	0.16437	45	-1.15	013.00	0
013	Fat (%)	2057	7.3200	0.12000	8.1562	0.79175	0.16437	45	-1.06	013.00	0
013	Fat (%)	0036	7.3340	0.01080	8.1562	0.79175	0.16437	45	-1.04	013.10	0
013	Fat (%)	0297	7.3600	0.06000	8.1562	0.79175	0.16437	45	-1.01	013.00	0
013	Fat (%)	0541	7.4350	0.03000	8.1562	0.79175	0.16437	45	-0.91	013.00	0
013	Fat (%)	0148	7.5050	0.29000	8.1562	0.79175	0.16437	45	-0.82	013.02	0
013	Fat (%)	0955	7.6800	0.10000	8.1562	0.79175	0.16437	45	-0.60	013.13	0
013	Fat (%)	0610	7.7000	0.20000	8.1562	0.79175	0.16437	45	-0.58	013.10	0
013	Fat (%)	0948	7.7550	0.07000	8.1562	0.79175	0.16437	45	-0.51	013.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
013	Fat (%)	0652	7.8000	0.00000	8.1562	0.79175	0.16437	45	-0.45	013.00	0
013	Fat (%)	0689	7.8000	0.20000	8.1562	0.79175	0.16437	45	-0.45	013.00	0
013	Fat (%)	0098	7.8600	0.10000	8.1562	0.79175	0.16437	45	-0.37	013.00	0
013	Fat (%)	0921	7.8700	0.00000	8.1562	0.79175	0.16437	45	-0.36	013.02	0
013	Fat (%)	2006	7.9050	0.51000	8.1562	0.79175	0.16437	45	-0.32	013.10	0
013	Fat (%)	0910	8.0450	0.11000	8.1562	0.79175	0.16437	45	-0.14	013.00	0
013	Fat (%)	0843	8.0650	0.39000	8.1562	0.79175	0.16437	45	-0.12	013.13	0
013	Fat (%)	0504	8.0950	0.13000	8.1562	0.79175	0.16437	45	-0.08	013.00	0
013	Fat (%)	0265	8.1500	0.30000	8.1562	0.79175	0.16437	45	-0.01	013.13	0
013	Fat (%)	0208	8.1900	0.84000	8.1562	0.79175	0.16437	45	0.04	013.02	0
013	Fat (%)	2004	8.2500	0.10000	8.1562	0.79175	0.16437	45	0.12	013.00	0
013	Fat (%)	0164	8.2600	0.16000	8.1562	0.79175	0.16437	45	0.13	013.02	0
013	Fat (%)	0353	8.2900	0.22000	8.1562	0.79175	0.16437	45	0.17	013.10	0
013	Fat (%)	2150	8.3850	0.57000	8.1562	0.79175	0.16437	45	0.29	013.02	0
013	Fat (%)	0100	8.5100	0.00000	8.1562	0.79175	0.16437	45	0.45	013.02	0
013	Fat (%)	0139	8.5200	0.08000	8.1562	0.79175	0.16437	45	0.46	013.00	0
013	Fat (%)	0309	8.5903	0.23580	8.1562	0.79175	0.16437	45	0.55	013.00	0
013	Fat (%)	0870	8.6465	0.08310	8.1562	0.79175	0.16437	45	0.62	013.02	0
013	Fat (%)	0874	8.6950	0.03000	8.1562	0.79175	0.16437	45	0.68	013.02	0
013	Fat (%)	0676	8.7050	0.03000	8.1562	0.79175	0.16437	45	0.69	013.02	0
013	Fat (%)	0861	8.7150	0.23000	8.1562	0.79175	0.16437	45	0.71	013.02	0
013	Fat (%)	0229	8.7400	0.04000	8.1562	0.79175	0.16437	45	0.74	013.00	0
013	Fat (%)	0650	8.7750	0.01000	8.1562	0.79175	0.16437	45	0.78	013.02	0
013	Fat (%)	0618	8.8224	0.06880	8.1562	0.79175	0.16437	45	0.84	013.00	0
013	Fat (%)	0354	8.9500	0.20000	8.1562	0.79175	0.16437	45	1.00	013.02	0
013	Fat (%)	0171	9.1200	0.08000	8.1562	0.79175	0.16437	45	1.22	013.02	0
013	Fat (%)	0553	9.2350	0.49000	8.1562	0.79175	0.16437	45	1.36	013.02	0
013	Fat (%)	0202	9.2500	0.00000	8.1562	0.79175	0.16437	45	1.38	013.00	0
013	Fat (%)	0675	9.2900	0.32000	8.1562	0.79175	0.16437	45	1.43	013.02	0
013	Fat (%)	0615	9.4600	0.52000	8.1562	0.79175	0.16437	45	1.65	013.00	0
013	Fat (%)	0682	9.6000	0.00000	8.1562	0.79175	0.16437	45	1.82	013.02	0
013	Fat (%)	0003	7.9450	1.5700	8.1562	0.79175	0.16437	45	-0.27	013.13	1
015	Aluminum (mg / kg (ppm))	0186	95.500	13.000	194.62	47.023	8.9089	12	-2.11	015.52	0
015	Aluminum (mg / kg (ppm))	0918	139.61	2.2270	194.62	47.023	8.9089	12	-1.17	015.43	0
015	Aluminum (mg / kg (ppm))	0297	160.50	11.000	194.62	47.023	8.9089	12	-0.73	015.43	0
015	Aluminum (mg / kg (ppm))	0049	175.46	11.180	194.62	47.023	8.9089	12	-0.41	015.41	0
015	Aluminum (mg / kg (ppm))	0560	182.20	7.4000	194.62	47.023	8.9089	12	-0.26	015.42	0
015	Aluminum (mg / kg (ppm))	0510	185.50	3.0000	194.62	47.023	8.9089	12	-0.19	015.43	0
015	Aluminum (mg / kg (ppm))	0345	205.95	6.1000	194.62	47.023	8.9089	12	0.24	015.43	0
015	Aluminum (mg / kg (ppm))	0520	209.50	3.0000	194.62	47.023	8.9089	12	0.32	015.41	0
015	Aluminum (mg / kg (ppm))	0553	216.50	3.0000	194.62	47.023	8.9089	12	0.47	015.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
015	Aluminum (mg / kg (ppm))	0353	234.50	11.000	194.62	47.023	8.9089	12	0.85	015.43	0
015	Aluminum (mg / kg (ppm))	0164	236.50	5.0000	194.62	47.023	8.9089	12	0.89	015.41	0
015	Aluminum (mg / kg (ppm))	0171	439.50	31.000	194.62	47.023	8.9089	12	5.21	015.41	0
015	Aluminum (mg / kg (ppm))	0169	0.00000	0.00000	194.62	47.023	8.9089	12		015.43	4
017	Boron (mg / kg (ppm))	0297	6.0000	0.00000	11.770	0.63753	0.65484	13	-9.05	017.43	0
017	Boron (mg / kg (ppm))	0560	10.965	0.31000	11.770	0.63753	0.65484	13	-1.26	017.42	0
017	Boron (mg / kg (ppm))	0226	11.130	0.04000	11.770	0.63753	0.65484	13	-1.00	017.41	0
017	Boron (mg / kg (ppm))	0353	11.500	3.0000	11.770	0.63753	0.65484	13	-0.42	017.43	0
017	Boron (mg / kg (ppm))	0045	11.650	2.3000	11.770	0.63753	0.65484	13	-0.19	017.42	0
017	Boron (mg / kg (ppm))	0693	11.695	0.18290	11.770	0.63753	0.65484	13	-0.12	017.42	0
017	Boron (mg / kg (ppm))	0918	11.770	0.12000	11.770	0.63753	0.65484	13	0.00	017.43	0
017	Boron (mg / kg (ppm))	0345	11.950	0.10000	11.770	0.63753	0.65484	13	0.28	017.43	0
017	Boron (mg / kg (ppm))	2053	11.985	0.01000	11.770	0.63753	0.65484	13	0.34	017.42	0
017	Boron (mg / kg (ppm))	0049	12.150	1.7600	11.770	0.63753	0.65484	13	0.60	017.41	0
017	Boron (mg / kg (ppm))	0358	12.325	0.37000	11.770	0.63753	0.65484	13	0.87	017.41	0
017	Boron (mg / kg (ppm))	0510	12.350	0.10000	11.770	0.63753	0.65484	13	0.91	017.43	0
017	Boron (mg / kg (ppm))	0294	13.410	0.22000	11.770	0.63753	0.65484	13	2.57	017.42	0
019	Calcium (%)	2062	0.40025	0.01490	1.0868	0.06894	0.03028	142	-9.96	019.00	0
019	Calcium (%)	2105	0.58000	0.08000	1.0868	0.06894	0.03028	142	-7.35	019.99	0
019	Calcium (%)	0529	0.86500	0.01000	1.0868	0.06894	0.03028	142	-3.22	019.31	0
019	Calcium (%)	0026	0.88500	0.03000	1.0868	0.06894	0.03028	142	-2.93	019.42	0
019	Calcium (%)	0552	0.96000	0.00000	1.0868	0.06894	0.03028	142	-1.84	019.00	0
019	Calcium (%)	0152	0.96500	0.03000	1.0868	0.06894	0.03028	142	-1.77	019.31	0
019	Calcium (%)	0154	0.96990	0.02640	1.0868	0.06894	0.03028	142	-1.70	019.52	0
019	Calcium (%)	0202	0.97500	0.07000	1.0868	0.06894	0.03028	142	-1.62	019.43	0
019	Calcium (%)	0358	0.98500	0.01000	1.0868	0.06894	0.03028	142	-1.48	019.41	0
019	Calcium (%)	0175	0.99000	0.02000	1.0868	0.06894	0.03028	142	-1.40	019.31	0
019	Calcium (%)	0874	0.99250	0.01300	1.0868	0.06894	0.03028	142	-1.37	019.31	0
019	Calcium (%)	0144	1.0000	0.02000	1.0868	0.06894	0.03028	142	-1.26	019.41	0
019	Calcium (%)	0510	1.0000	0.02000	1.0868	0.06894	0.03028	142	-1.26	019.43	0
019	Calcium (%)	0233	1.0050	0.05000	1.0868	0.06894	0.03028	142	-1.19	019.31	0
019	Calcium (%)	0028	1.0050	0.01000	1.0868	0.06894	0.03028	142	-1.19	019.43	0
019	Calcium (%)	0033	1.0050	0.03000	1.0868	0.06894	0.03028	142	-1.19	019.43	0
019	Calcium (%)	0622	1.0106	0.00010	1.0868	0.06894	0.03028	142	-1.11	019.00	0
019	Calcium (%)	0903	1.0150	0.01000	1.0868	0.06894	0.03028	142	-1.04	019.00	0
019	Calcium (%)	0563	1.0200	0.05370	1.0868	0.06894	0.03028	142	-0.97	019.31	0
019	Calcium (%)	0047	1.0200	0.02000	1.0868	0.06894	0.03028	142	-0.97	019.52	0
019	Calcium (%)	0948	1.0220	0.01400	1.0868	0.06894	0.03028	142	-0.94	019.31	0
019	Calcium (%)	0620	1.0248	0.10070	1.0868	0.06894	0.03028	142	-0.90	019.00	0
019	Calcium (%)	0897	1.0250	0.01000	1.0868	0.06894	0.03028	142	-0.90	019.00	0
019	Calcium (%)	2093	1.0250	0.01000	1.0868	0.06894	0.03028	142	-0.90	019.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	0098	1.0250	0.03000	1.0868	0.06894	0.03028	142	-0.90	019.41	0
019	Calcium (%)	0910	1.0250	0.05000	1.0868	0.06894	0.03028	142	-0.90	019.41	0
019	Calcium (%)	0142	1.0292	0.02190	1.0868	0.06894	0.03028	142	-0.84	019.31	0
019	Calcium (%)	0893	1.0300	0.00000	1.0868	0.06894	0.03028	142	-0.82	019.00	0
019	Calcium (%)	2109	1.0305	0.00100	1.0868	0.06894	0.03028	142	-0.82	019.00	0
019	Calcium (%)	0693	1.0349	0.02790	1.0868	0.06894	0.03028	142	-0.75	019.42	0
019	Calcium (%)	0511	1.0350	0.01000	1.0868	0.06894	0.03028	142	-0.75	019.41	0
019	Calcium (%)	0001	1.0354	0.07330	1.0868	0.06894	0.03028	142	-0.75	019.31	0
019	Calcium (%)	0623	1.0365	0.07300	1.0868	0.06894	0.03028	142	-0.73	019.00	0
019	Calcium (%)	0560	1.0392	0.02660	1.0868	0.06894	0.03028	142	-0.69	019.42	0
019	Calcium (%)	2068	1.0400	0.12000	1.0868	0.06894	0.03028	142	-0.68	019.00	0
019	Calcium (%)	0208	1.0440	0.00600	1.0868	0.06894	0.03028	142	-0.62	019.41	0
019	Calcium (%)	0651	1.0450	0.03000	1.0868	0.06894	0.03028	142	-0.61	019.00	0
019	Calcium (%)	0265	1.0450	0.01000	1.0868	0.06894	0.03028	142	-0.61	019.42	0
019	Calcium (%)	0297	1.0450	0.15000	1.0868	0.06894	0.03028	142	-0.61	019.43	0
019	Calcium (%)	0675	1.0450	0.03000	1.0868	0.06894	0.03028	142	-0.61	019.43	0
019	Calcium (%)	0186	1.0450	0.01000	1.0868	0.06894	0.03028	142	-0.61	019.52	0
019	Calcium (%)	0263	1.0467	0.00130	1.0868	0.06894	0.03028	142	-0.58	019.32	0
019	Calcium (%)	0520	1.0500	0.02000	1.0868	0.06894	0.03028	142	-0.53	019.41	0
019	Calcium (%)	0848	1.0500	0.00000	1.0868	0.06894	0.03028	142	-0.53	019.41	0
019	Calcium (%)	0878	1.0500	0.10000	1.0868	0.06894	0.03028	142	-0.53	019.41	0
019	Calcium (%)	0043	1.0500	0.00000	1.0868	0.06894	0.03028	142	-0.53	019.43	0
019	Calcium (%)	0300	1.0550	0.09000	1.0868	0.06894	0.03028	142	-0.46	019.33	0
019	Calcium (%)	0205	1.0550	0.05000	1.0868	0.06894	0.03028	142	-0.46	019.42	0
019	Calcium (%)	0610	1.0550	0.03000	1.0868	0.06894	0.03028	142	-0.46	019.43	0
019	Calcium (%)	0036	1.0552	0.02110	1.0868	0.06894	0.03028	142	-0.46	019.43	0
019	Calcium (%)	2006	1.0560	0.02000	1.0868	0.06894	0.03028	142	-0.45	019.09	0
019	Calcium (%)	0298	1.0600	0.02000	1.0868	0.06894	0.03028	142	-0.39	019.41	0
019	Calcium (%)	2012	1.0600	0.00000	1.0868	0.06894	0.03028	142	-0.39	019.41	0
019	Calcium (%)	0045	1.0600	0.02000	1.0868	0.06894	0.03028	142	-0.39	019.42	0
019	Calcium (%)	0629	1.0600	0.00000	1.0868	0.06894	0.03028	142	-0.39	019.43	0
019	Calcium (%)	2004	1.0600	0.04000	1.0868	0.06894	0.03028	142	-0.39	019.44	0
019	Calcium (%)	0895	1.0650	0.01000	1.0868	0.06894	0.03028	142	-0.32	019.00	0
019	Calcium (%)	0849	1.0650	0.01000	1.0868	0.06894	0.03028	142	-0.32	019.08	0
019	Calcium (%)	0939	1.0650	0.01000	1.0868	0.06894	0.03028	142	-0.32	019.31	0
019	Calcium (%)	0921	1.0660	0.00000	1.0868	0.06894	0.03028	142	-0.30	019.31	0
019	Calcium (%)	0014	1.0700	0.04000	1.0868	0.06894	0.03028	142	-0.24	019.31	0
019	Calcium (%)	0139	1.0700	0.00000	1.0868	0.06894	0.03028	142	-0.24	019.41	0
019	Calcium (%)	0013	1.0700	0.08000	1.0868	0.06894	0.03028	142	-0.24	019.43	0
019	Calcium (%)	0164	1.0750	0.03000	1.0868	0.06894	0.03028	142	-0.17	019.41	0
019	Calcium (%)	2009	1.0767	0.01090	1.0868	0.06894	0.03028	142	-0.15	019.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	2097	1.0791	0.00560	1.0868	0.06894	0.03028	142	-0.11	019.31	0
019	Calcium (%)	0940	1.0800	0.08000	1.0868	0.06894	0.03028	142	-0.10	019.08	0
019	Calcium (%)	0354	1.0800	0.00000	1.0868	0.06894	0.03028	142	-0.10	019.31	0
019	Calcium (%)	2022	1.0800	0.00000	1.0868	0.06894	0.03028	142	-0.10	019.31	0
019	Calcium (%)	0148	1.0800	0.02000	1.0868	0.06894	0.03028	142	-0.10	019.41	0
019	Calcium (%)	2152	1.0800	0.00000	1.0868	0.06894	0.03028	142	-0.10	019.44	0
019	Calcium (%)	2006	1.0805	0.00500	1.0868	0.06894	0.03028	142	-0.09	019.00	0
019	Calcium (%)	0653	1.0810	0.00600	1.0868	0.06894	0.03028	142	-0.08	019.31	0
019	Calcium (%)	0881	1.0815	0.01500	1.0868	0.06894	0.03028	142	-0.08	019.00	0
019	Calcium (%)	0505	1.0850	0.01000	1.0868	0.06894	0.03028	142	-0.03	019.31	0
019	Calcium (%)	0034	1.0883	0.09190	1.0868	0.06894	0.03028	142	0.02	019.43	0
019	Calcium (%)	0186	1.0900	0.00000	1.0868	0.06894	0.03028	142	0.05	019.42	0
019	Calcium (%)	0019	1.0950	0.03000	1.0868	0.06894	0.03028	142	0.12	019.41	0
019	Calcium (%)	2145	1.0950	0.01000	1.0868	0.06894	0.03028	142	0.12	019.41	0
019	Calcium (%)	0964	1.0965	0.05500	1.0868	0.06894	0.03028	142	0.14	019.43	0
019	Calcium (%)	0208	1.0975	0.02900	1.0868	0.06894	0.03028	142	0.16	019.31	0
019	Calcium (%)	0918	1.0975	0.04300	1.0868	0.06894	0.03028	142	0.16	019.43	0
019	Calcium (%)	0043	1.1000	0.00000	1.0868	0.06894	0.03028	142	0.19	019.00	0
019	Calcium (%)	0878	1.1000	0.02000	1.0868	0.06894	0.03028	142	0.19	019.00	0
019	Calcium (%)	0074	1.1000	0.02000	1.0868	0.06894	0.03028	142	0.19	019.41	0
019	Calcium (%)	0345	1.1000	0.02000	1.0868	0.06894	0.03028	142	0.19	019.43	0
019	Calcium (%)	0955	1.1000	0.02000	1.0868	0.06894	0.03028	142	0.19	019.44	0
019	Calcium (%)	0870	1.1035	0.00900	1.0868	0.06894	0.03028	142	0.24	019.42	0
019	Calcium (%)	0885	1.1049	0.00590	1.0868	0.06894	0.03028	142	0.26	019.08	0
019	Calcium (%)	2129	1.1050	0.05000	1.0868	0.06894	0.03028	142	0.26	019.42	0
019	Calcium (%)	0083	1.1050	0.03000	1.0868	0.06894	0.03028	142	0.26	019.43	0
019	Calcium (%)	0612	1.1090	0.01000	1.0868	0.06894	0.03028	142	0.32	019.32	0
019	Calcium (%)	0687	1.1100	0.08000	1.0868	0.06894	0.03028	142	0.34	019.31	0
019	Calcium (%)	0194	1.1100	0.02000	1.0868	0.06894	0.03028	142	0.34	019.00	0
019	Calcium (%)	2068	1.1100	0.00000	1.0868	0.06894	0.03028	142	0.34	019.31	0
019	Calcium (%)	0294	1.1150	0.01000	1.0868	0.06894	0.03028	142	0.41	019.42	0
019	Calcium (%)	0366	1.1150	0.01000	1.0868	0.06894	0.03028	142	0.41	019.42	0
019	Calcium (%)	0049	1.1200	0.02000	1.0868	0.06894	0.03028	142	0.48	019.41	0
019	Calcium (%)	0425	1.1200	0.00000	1.0868	0.06894	0.03028	142	0.48	019.43	0
019	Calcium (%)	2089	1.1200	0.00000	1.0868	0.06894	0.03028	142	0.48	019.43	0
019	Calcium (%)	0504	1.1235	0.06900	1.0868	0.06894	0.03028	142	0.53	019.42	0
019	Calcium (%)	0683	1.1250	0.07000	1.0868	0.06894	0.03028	142	0.55	019.00	0
019	Calcium (%)	0353	1.1250	0.05000	1.0868	0.06894	0.03028	142	0.55	019.43	0
019	Calcium (%)	2146	1.1250	0.01800	1.0868	0.06894	0.03028	142	0.55	019.51	0
019	Calcium (%)	0619	1.1300	0.02000	1.0868	0.06894	0.03028	142	0.63	019.41	0
019	Calcium (%)	0682	1.1300	0.00000	1.0868	0.06894	0.03028	142	0.63	019.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	0861	1.1300	0.02000	1.0868	0.06894	0.03028	142	0.63	019.43	0
019	Calcium (%)	0006	1.1345	0.01500	1.0868	0.06894	0.03028	142	0.69	019.43	0
019	Calcium (%)	0035	1.1350	0.11000	1.0868	0.06894	0.03028	142	0.70	019.99	0
019	Calcium (%)	0098	1.1400	0.04000	1.0868	0.06894	0.03028	142	0.77	019.43	0
019	Calcium (%)	0504	1.1415	0.02100	1.0868	0.06894	0.03028	142	0.79	019.32	0
019	Calcium (%)	0009	1.1426	0.00500	1.0868	0.06894	0.03028	142	0.81	019.42	0
019	Calcium (%)	0968	1.1435	0.00500	1.0868	0.06894	0.03028	142	0.82	019.43	0
019	Calcium (%)	0650	1.1450	0.05000	1.0868	0.06894	0.03028	142	0.84	019.31	0
019	Calcium (%)	0598	1.1495	0.02300	1.0868	0.06894	0.03028	142	0.91	019.41	0
019	Calcium (%)	0413	1.1500	0.10000	1.0868	0.06894	0.03028	142	0.92	019.41	0
019	Calcium (%)	0035	1.1500	0.04000	1.0868	0.06894	0.03028	142	0.92	019.42	0
019	Calcium (%)	0036	1.1512	0.01450	1.0868	0.06894	0.03028	142	0.93	019.03	0
019	Calcium (%)	0553	1.1550	0.01000	1.0868	0.06894	0.03028	142	0.99	019.53	0
019	Calcium (%)	0169	1.1650	0.01000	1.0868	0.06894	0.03028	142	1.13	019.32	0
019	Calcium (%)	0278	1.1650	0.01000	1.0868	0.06894	0.03028	142	1.13	019.42	0
019	Calcium (%)	0035	1.1700	0.02000	1.0868	0.06894	0.03028	142	1.21	019.32	0
019	Calcium (%)	0035	1.1700	0.02000	1.0868	0.06894	0.03028	142	1.21	019.43	0
019	Calcium (%)	0572	1.1700	0.12000	1.0868	0.06894	0.03028	142	1.21	019.53	0
019	Calcium (%)	0100	1.1700	0.00000	1.0868	0.06894	0.03028	142	1.21	019.99	0
019	Calcium (%)	0027	1.1720	0.05600	1.0868	0.06894	0.03028	142	1.24	019.43	0
019	Calcium (%)	0407	1.1763	0.03930	1.0868	0.06894	0.03028	142	1.30	019.41	0
019	Calcium (%)	0037	1.1800	0.10000	1.0868	0.06894	0.03028	142	1.35	019.42	0
019	Calcium (%)	0689	1.1850	0.03000	1.0868	0.06894	0.03028	142	1.42	019.31	0
019	Calcium (%)	0003	1.1950	0.01000	1.0868	0.06894	0.03028	142	1.57	019.41	0
019	Calcium (%)	0629	1.2050	0.01000	1.0868	0.06894	0.03028	142	1.71	019.08	0
019	Calcium (%)	0226	1.2050	0.03000	1.0868	0.06894	0.03028	142	1.71	019.41	0
019	Calcium (%)	0689	1.2100	0.06000	1.0868	0.06894	0.03028	142	1.79	019.08	0
019	Calcium (%)	0504	1.2125	0.08900	1.0868	0.06894	0.03028	142	1.82	019.33	0
019	Calcium (%)	0187	1.2200	0.00000	1.0868	0.06894	0.03028	142	1.93	019.42	0
019	Calcium (%)	0357	1.2233	0.05230	1.0868	0.06894	0.03028	142	1.98	019.42	0
019	Calcium (%)	0190	1.2250	0.01000	1.0868	0.06894	0.03028	142	2.01	019.42	0
019	Calcium (%)	0726	1.2250	0.00200	1.0868	0.06894	0.03028	142	2.01	019.42	0
019	Calcium (%)	0108	1.2450	0.17000	1.0868	0.06894	0.03028	142	2.30	019.31	0
019	Calcium (%)	0590	1.2500	0.06000	1.0868	0.06894	0.03028	142	2.37	019.08	0
019	Calcium (%)	0692	1.3000	0.06000	1.0868	0.06894	0.03028	142	3.09	019.42	0
019	Calcium (%)	0405	1.4750	0.01000	1.0868	0.06894	0.03028	142	5.63	019.41	0
019	Calcium (%)	2055	1.0115	0.18970	1.0868	0.06894	0.03028	142	-1.09	019.53	1
019	Calcium (%)	0042	1.1900	0.38000	1.0868	0.06894	0.03028	142	1.50	019.42	1
019	Calcium (%)	0536	5.5500	0.14000	1.0868	0.06894	0.03028	142	64.74	019.31	2
019	Calcium (%)	0171	6.4950	0.05000	1.0868	0.06894	0.03028	142	78.45	019.41	2
021	Cobalt (mg / kg (ppm))	0171	5.5000	0.00000	6.4468	0.56898	0.19046	16	-1.66	021.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
021	Cobalt (mg / kg (ppm))	0345	5.7100	0.10000	6.4468	0.56898	0.19046	16	-1.29	021.43	0
021	Cobalt (mg / kg (ppm))	0693	5.7602	0.48210	6.4468	0.56898	0.19046	16	-1.21	021.42	0
021	Cobalt (mg / kg (ppm))	0510	6.2650	0.01000	6.4468	0.56898	0.19046	16	-0.32	021.43	0
021	Cobalt (mg / kg (ppm))	0098	6.2935	0.40300	6.4468	0.56898	0.19046	16	-0.27	021.53	0
021	Cobalt (mg / kg (ppm))	0043	6.3350	0.07000	6.4468	0.56898	0.19046	16	-0.20	021.43	0
021	Cobalt (mg / kg (ppm))	0164	6.3500	0.10000	6.4468	0.56898	0.19046	16	-0.17	021.31	0
021	Cobalt (mg / kg (ppm))	0186	6.3500	0.30000	6.4468	0.56898	0.19046	16	-0.17	021.52	0
021	Cobalt (mg / kg (ppm))	0918	6.4515	0.33220	6.4468	0.56898	0.19046	16	0.01	021.53	0
021	Cobalt (mg / kg (ppm))	0964	6.4900	0.00000	6.4468	0.56898	0.19046	16	0.08	021.43	0
021	Cobalt (mg / kg (ppm))	0560	6.5750	0.01000	6.4468	0.56898	0.19046	16	0.23	021.42	0
021	Cobalt (mg / kg (ppm))	2053	6.5800	0.20000	6.4468	0.56898	0.19046	16	0.23	021.42	0
021	Cobalt (mg / kg (ppm))	0910	6.7450	0.15000	6.4468	0.56898	0.19046	16	0.52	021.52	0
021	Cobalt (mg / kg (ppm))	0939	7.0500	0.28000	6.4468	0.56898	0.19046	16	1.06	021.31	0
021	Cobalt (mg / kg (ppm))	0553	7.6400	0.48000	6.4468	0.56898	0.19046	16	2.10	021.53	0
021	Cobalt (mg / kg (ppm))	0590	7.7450	0.13000	6.4468	0.56898	0.19046	16	2.28	021.99	0
021	Cobalt (mg / kg (ppm))	2055	6.7815	1.0650	6.4468	0.56898	0.19046	16	0.59	021.53	1
021	Cobalt (mg / kg (ppm))	0169	0.00000	0.00000	6.4468	0.56898	0.19046	16		021.43	4
022	Copper (mg / kg (ppm))	0026	55.000	0.00000	91.097	6.7122	2.7857	95	-5.38	022.42	0
022	Copper (mg / kg (ppm))	0689	60.900	4.4000	91.097	6.7122	2.7857	95	-4.50	022.31	0
022	Copper (mg / kg (ppm))	2055	74.500	9.0000	91.097	6.7122	2.7857	95	-2.47	022.53	0
022	Copper (mg / kg (ppm))	0033	78.300	5.0000	91.097	6.7122	2.7857	95	-1.91	022.43	0
022	Copper (mg / kg (ppm))	0955	80.700	2.0000	91.097	6.7122	2.7857	95	-1.55	022.44	0
022	Copper (mg / kg (ppm))	0921	80.800	0.00000	91.097	6.7122	2.7857	95	-1.53	022.31	0
022	Copper (mg / kg (ppm))	0878	81.500	5.0000	91.097	6.7122	2.7857	95	-1.43	022.41	0
022	Copper (mg / kg (ppm))	0003	82.000	0.00000	91.097	6.7122	2.7857	95	-1.36	022.41	0
022	Copper (mg / kg (ppm))	0098	82.015	12.650	91.097	6.7122	2.7857	95	-1.35	022.41	0
022	Copper (mg / kg (ppm))	0910	82.500	5.0000	91.097	6.7122	2.7857	95	-1.28	022.41	0
022	Copper (mg / kg (ppm))	0619	83.500	1.2000	91.097	6.7122	2.7857	95	-1.13	022.41	0
022	Copper (mg / kg (ppm))	0148	84.000	4.0000	91.097	6.7122	2.7857	95	-1.06	022.41	0
022	Copper (mg / kg (ppm))	0529	84.190	1.1200	91.097	6.7122	2.7857	95	-1.03	022.31	0
022	Copper (mg / kg (ppm))	0001	84.290	0.74000	91.097	6.7122	2.7857	95	-1.01	022.31	0
022	Copper (mg / kg (ppm))	0208	84.850	1.1000	91.097	6.7122	2.7857	95	-0.93	022.31	0
022	Copper (mg / kg (ppm))	0043	85.050	0.90000	91.097	6.7122	2.7857	95	-0.90	022.43	0
022	Copper (mg / kg (ppm))	0510	85.500	1.0000	91.097	6.7122	2.7857	95	-0.83	022.43	0
022	Copper (mg / kg (ppm))	2012	85.710	0.82000	91.097	6.7122	2.7857	95	-0.80	022.41	0
022	Copper (mg / kg (ppm))	0629	85.900	0.00000	91.097	6.7122	2.7857	95	-0.77	022.43	0
022	Copper (mg / kg (ppm))	0175	86.000	4.0000	91.097	6.7122	2.7857	95	-0.76	022.31	0
022	Copper (mg / kg (ppm))	0164	86.000	2.0000	91.097	6.7122	2.7857	95	-0.76	022.41	0
022	Copper (mg / kg (ppm))	0208	86.085	3.1300	91.097	6.7122	2.7857	95	-0.75	022.41	0
022	Copper (mg / kg (ppm))	0202	86.185	3.7500	91.097	6.7122	2.7857	95	-0.73	022.43	0
022	Copper (mg / kg (ppm))	0037	86.450	13.900	91.097	6.7122	2.7857	95	-0.69	022.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
022	Copper (mg / kg (ppm))	0939	86.550	2.8600	91.097	6.7122	2.7857	95	-0.68	022.31	0
022	Copper (mg / kg (ppm))	0598	86.720	1.2200	91.097	6.7122	2.7857	95	-0.65	022.41	0
022	Copper (mg / kg (ppm))	2068	87.035	2.5280	91.097	6.7122	2.7857	95	-0.61	022.31	0
022	Copper (mg / kg (ppm))	0294	87.420	0.24000	91.097	6.7122	2.7857	95	-0.55	022.42	0
022	Copper (mg / kg (ppm))	0511	87.500	1.0000	91.097	6.7122	2.7857	95	-0.54	022.41	0
022	Copper (mg / kg (ppm))	0629	88.000	0.00000	91.097	6.7122	2.7857	95	-0.46	022.31	0
022	Copper (mg / kg (ppm))	0505	88.500	2.0000	91.097	6.7122	2.7857	95	-0.39	022.31	0
022	Copper (mg / kg (ppm))	0366	88.500	3.0000	91.097	6.7122	2.7857	95	-0.39	022.42	0
022	Copper (mg / kg (ppm))	0345	88.550	2.1000	91.097	6.7122	2.7857	95	-0.38	022.43	0
022	Copper (mg / kg (ppm))	0049	88.830	12.500	91.097	6.7122	2.7857	95	-0.34	022.41	0
022	Copper (mg / kg (ppm))	0205	88.950	2.7000	91.097	6.7122	2.7857	95	-0.32	022.42	0
022	Copper (mg / kg (ppm))	0098	88.960	4.3200	91.097	6.7122	2.7857	95	-0.32	022.43	0
022	Copper (mg / kg (ppm))	0520	89.000	0.00000	91.097	6.7122	2.7857	95	-0.31	022.41	0
022	Copper (mg / kg (ppm))	0297	89.000	6.0000	91.097	6.7122	2.7857	95	-0.31	022.43	0
022	Copper (mg / kg (ppm))	0504	89.165	0.57000	91.097	6.7122	2.7857	95	-0.29	022.33	0
022	Copper (mg / kg (ppm))	0590	89.200	0.20000	91.097	6.7122	2.7857	95	-0.28	022.99	0
022	Copper (mg / kg (ppm))	0186	89.500	3.0000	91.097	6.7122	2.7857	95	-0.24	022.52	0
022	Copper (mg / kg (ppm))	0100	89.500	5.0000	91.097	6.7122	2.7857	95	-0.24	022.99	0
022	Copper (mg / kg (ppm))	0354	89.975	0.75000	91.097	6.7122	2.7857	95	-0.17	022.31	0
022	Copper (mg / kg (ppm))	0278	90.150	1.3000	91.097	6.7122	2.7857	95	-0.14	022.42	0
022	Copper (mg / kg (ppm))	0675	90.180	3.1200	91.097	6.7122	2.7857	95	-0.14	022.43	0
022	Copper (mg / kg (ppm))	0653	90.365	3.4100	91.097	6.7122	2.7857	95	-0.11	022.31	0
022	Copper (mg / kg (ppm))	0265	90.500	5.0000	91.097	6.7122	2.7857	95	-0.09	022.42	0
022	Copper (mg / kg (ppm))	2022	91.000	2.0000	91.097	6.7122	2.7857	95	-0.01	022.31	0
022	Copper (mg / kg (ppm))	0610	91.000	6.0000	91.097	6.7122	2.7857	95	-0.01	022.43	0
022	Copper (mg / kg (ppm))	0407	91.064	5.7784	91.097	6.7122	2.7857	95	0.00	022.41	0
022	Copper (mg / kg (ppm))	0013	91.350	0.10000	91.097	6.7122	2.7857	95	0.04	022.43	0
022	Copper (mg / kg (ppm))	0553	91.450	6.7000	91.097	6.7122	2.7857	95	0.05	022.53	0
022	Copper (mg / kg (ppm))	0918	91.495	2.3100	91.097	6.7122	2.7857	95	0.06	022.43	0
022	Copper (mg / kg (ppm))	0074	91.500	3.0000	91.097	6.7122	2.7857	95	0.06	022.41	0
022	Copper (mg / kg (ppm))	0035	92.000	2.0000	91.097	6.7122	2.7857	95	0.13	022.43	0
022	Copper (mg / kg (ppm))	0504	92.120	2.6000	91.097	6.7122	2.7857	95	0.15	022.32	0
022	Copper (mg / kg (ppm))	0083	92.500	3.0000	91.097	6.7122	2.7857	95	0.21	022.43	0
022	Copper (mg / kg (ppm))	0563	92.741	0.90260	91.097	6.7122	2.7857	95	0.24	022.31	0
022	Copper (mg / kg (ppm))	0019	92.850	2.6000	91.097	6.7122	2.7857	95	0.26	022.41	0
022	Copper (mg / kg (ppm))	0572	92.850	2.9000	91.097	6.7122	2.7857	95	0.26	022.53	0
022	Copper (mg / kg (ppm))	0870	93.050	3.7600	91.097	6.7122	2.7857	95	0.29	022.42	0
022	Copper (mg / kg (ppm))	2149	93.175	1.9300	91.097	6.7122	2.7857	95	0.31	022.31	0
022	Copper (mg / kg (ppm))	0413	93.250	0.50000	91.097	6.7122	2.7857	95	0.32	022.42	0
022	Copper (mg / kg (ppm))	0682	93.400	0.00000	91.097	6.7122	2.7857	95	0.34	022.43	0
022	Copper (mg / kg (ppm))	0047	93.400	2.7000	91.097	6.7122	2.7857	95	0.34	022.52	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
022	Copper (mg / kg (ppm))	0045	94.100	0.20000	91.097	6.7122	2.7857	95	0.45	022.42	0
022	Copper (mg / kg (ppm))	0861	94.400	3.4000	91.097	6.7122	2.7857	95	0.49	022.43	0
022	Copper (mg / kg (ppm))	0964	94.515	8.2100	91.097	6.7122	2.7857	95	0.51	022.43	0
022	Copper (mg / kg (ppm))	0187	94.645	0.09000	91.097	6.7122	2.7857	95	0.53	022.42	0
022	Copper (mg / kg (ppm))	0405	95.150	0.10000	91.097	6.7122	2.7857	95	0.60	022.41	0
022	Copper (mg / kg (ppm))	0035	95.500	3.0000	91.097	6.7122	2.7857	95	0.66	022.42	0
022	Copper (mg / kg (ppm))	0968	95.500	1.0000	91.097	6.7122	2.7857	95	0.66	022.43	0
022	Copper (mg / kg (ppm))	0425	95.850	0.10000	91.097	6.7122	2.7857	95	0.71	022.43	0
022	Copper (mg / kg (ppm))	0948	96.286	0.31600	91.097	6.7122	2.7857	95	0.77	022.33	0
022	Copper (mg / kg (ppm))	2004	96.400	0.60000	91.097	6.7122	2.7857	95	0.79	022.44	0
022	Copper (mg / kg (ppm))	2129	96.730	2.7000	91.097	6.7122	2.7857	95	0.84	022.99	0
022	Copper (mg / kg (ppm))	0035	97.000	2.0000	91.097	6.7122	2.7857	95	0.88	022.32	0
022	Copper (mg / kg (ppm))	0693	97.105	7.3828	91.097	6.7122	2.7857	95	0.90	022.42	0
022	Copper (mg / kg (ppm))	0685	97.380	0.20000	91.097	6.7122	2.7857	95	0.94	022.42	0
022	Copper (mg / kg (ppm))	0035	97.500	1.0000	91.097	6.7122	2.7857	95	0.95	022.33	0
022	Copper (mg / kg (ppm))	0560	97.950	0.50000	91.097	6.7122	2.7857	95	1.02	022.42	0
022	Copper (mg / kg (ppm))	0027	98.679	0.79400	91.097	6.7122	2.7857	95	1.13	022.43	0
022	Copper (mg / kg (ppm))	0358	98.735	2.6900	91.097	6.7122	2.7857	95	1.14	022.41	0
022	Copper (mg / kg (ppm))	0010	99.000	2.0000	91.097	6.7122	2.7857	95	1.18	022.33	0
022	Copper (mg / kg (ppm))	0190	100.41	6.7800	91.097	6.7122	2.7857	95	1.39	022.42	0
022	Copper (mg / kg (ppm))	0357	100.85	0.89200	91.097	6.7122	2.7857	95	1.45	022.42	0
022	Copper (mg / kg (ppm))	0849	101.06	0.52000	91.097	6.7122	2.7857	95	1.48	022.31	0
022	Copper (mg / kg (ppm))	2089	101.17	5.0600	91.097	6.7122	2.7857	95	1.50	022.43	0
022	Copper (mg / kg (ppm))	2053	102.05	3.7000	91.097	6.7122	2.7857	95	1.63	022.42	0
022	Copper (mg / kg (ppm))	0726	102.20	0.00000	91.097	6.7122	2.7857	95	1.65	022.42	0
022	Copper (mg / kg (ppm))	0226	103.00	4.0000	91.097	6.7122	2.7857	95	1.77	022.41	0
022	Copper (mg / kg (ppm))	0353	103.50	7.0000	91.097	6.7122	2.7857	95	1.85	022.43	0
022	Copper (mg / kg (ppm))	2146	103.95	5.1000	91.097	6.7122	2.7857	95	1.91	022.51	0
022	Copper (mg / kg (ppm))	0186	104.00	2.0000	91.097	6.7122	2.7857	95	1.92	022.42	0
022	Copper (mg / kg (ppm))	0171	451.50	1.0000	91.097	6.7122	2.7857	95	53.69	022.41	0
022	Copper (mg / kg (ppm))	0042	99.850	18.300	91.097	6.7122	2.7857	95	1.30	022.42	1
022	Copper (mg / kg (ppm))	0692	141.50	27.000	91.097	6.7122	2.7857	95	7.51	022.42	1
022	Copper (mg / kg (ppm))	0536	268.13	25.550	91.097	6.7122	2.7857	95	26.37	022.31	1
022	Copper (mg / kg (ppm))	0009	826.74	3.2990	91.097	6.7122	2.7857	95	109.60	022.42	2
022	Copper (mg / kg (ppm))	0169	0.00000	0.00000	91.097	6.7122	2.7857	95		022.43	4
023	Fluorine (mg / kg (ppm))	0910	15.650	0.70000				1		023.01	0
024	Iodine (mg / kg (ppm))	0186	17.500	7.0000				2		024.01	0
024	Iodine (mg / kg (ppm))	0186	31.000	20.000				2		024.52	0
025	Iron (mg / kg (ppm))	0187	124.53	6.6800	587.82	57.289	20.061	88	-8.09	025.42	0
025	Iron (mg / kg (ppm))	2055	235.50	15.000	587.82	57.289	20.061	88	-6.15	025.53	0
025	Iron (mg / kg (ppm))	0186	273.50	87.000	587.82	57.289	20.061	88	-5.49	025.52	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
025	Iron (mg / kg (ppm))	0726	316.55	4.5000	587.82	57.289	20.061	88	-4.74	025.42	0
025	Iron (mg / kg (ppm))	0186	360.50	71.000	587.82	57.289	20.061	88	-3.97	025.42	0
025	Iron (mg / kg (ppm))	0968	403.00	0.00000	587.82	57.289	20.061	88	-3.23	025.43	0
025	Iron (mg / kg (ppm))	0026	421.00	34.000	587.82	57.289	20.061	88	-2.91	025.42	0
025	Iron (mg / kg (ppm))	0154	426.57	16.570	587.82	57.289	20.061	88	-2.81	025.52	0
025	Iron (mg / kg (ppm))	0190	436.55	30.700	587.82	57.289	20.061	88	-2.64	025.42	0
025	Iron (mg / kg (ppm))	0921	443.00	0.00000	587.82	57.289	20.061	88	-2.53	025.31	0
025	Iron (mg / kg (ppm))	2145	463.00	2.0000	587.82	57.289	20.061	88	-2.18	025.41	0
025	Iron (mg / kg (ppm))	0682	505.04	0.00000	587.82	57.289	20.061	88	-1.44	025.43	0
025	Iron (mg / kg (ppm))	0047	511.47	2.5900	587.82	57.289	20.061	88	-1.33	025.52	0
025	Iron (mg / kg (ppm))	0955	512.50	29.000	587.82	57.289	20.061	88	-1.31	025.41	0
025	Iron (mg / kg (ppm))	0345	517.35	5.7000	587.82	57.289	20.061	88	-1.23	025.43	0
025	Iron (mg / kg (ppm))	0265	520.50	41.000	587.82	57.289	20.061	88	-1.18	025.42	0
025	Iron (mg / kg (ppm))	0619	527.50	7.0000	587.82	57.289	20.061	88	-1.05	025.41	0
025	Iron (mg / kg (ppm))	0405	539.25	1.7000	587.82	57.289	20.061	88	-0.85	025.41	0
025	Iron (mg / kg (ppm))	0598	542.95	4.9000	587.82	57.289	20.061	88	-0.78	025.41	0
025	Iron (mg / kg (ppm))	0629	559.00	6.0000	587.82	57.289	20.061	88	-0.50	025.31	0
025	Iron (mg / kg (ppm))	0629	561.00	0.00000	587.82	57.289	20.061	88	-0.47	025.43	0
025	Iron (mg / kg (ppm))	0278	562.50	9.0000	587.82	57.289	20.061	88	-0.44	025.42	0
025	Iron (mg / kg (ppm))	0425	563.15	9.1000	587.82	57.289	20.061	88	-0.43	025.43	0
025	Iron (mg / kg (ppm))	0028	563.50	31.000	587.82	57.289	20.061	88	-0.42	025.43	0
025	Iron (mg / kg (ppm))	2089	566.29	21.660	587.82	57.289	20.061	88	-0.38	025.43	0
025	Iron (mg / kg (ppm))	0878	566.50	5.0000	587.82	57.289	20.061	88	-0.37	025.41	0
025	Iron (mg / kg (ppm))	0035	569.50	19.000	587.82	57.289	20.061	88	-0.32	025.32	0
025	Iron (mg / kg (ppm))	0689	571.60	15.400	587.82	57.289	20.061	88	-0.28	025.31	0
025	Iron (mg / kg (ppm))	0013	573.50	11.000	587.82	57.289	20.061	88	-0.25	025.43	0
025	Iron (mg / kg (ppm))	0294	573.86	5.2800	587.82	57.289	20.061	88	-0.24	025.42	0
025	Iron (mg / kg (ppm))	2022	575.00	0.00000	587.82	57.289	20.061	88	-0.22	025.31	0
025	Iron (mg / kg (ppm))	0939	578.50	11.200	587.82	57.289	20.061	88	-0.16	025.31	0
025	Iron (mg / kg (ppm))	0035	578.50	33.000	587.82	57.289	20.061	88	-0.16	025.43	0
025	Iron (mg / kg (ppm))	0510	579.00	0.00000	587.82	57.289	20.061	88	-0.15	025.43	0
025	Iron (mg / kg (ppm))	0590	579.00	0.00000	587.82	57.289	20.061	88	-0.15	025.99	0
025	Iron (mg / kg (ppm))	0001	579.25	21.500	587.82	57.289	20.061	88	-0.15	025.34	0
025	Iron (mg / kg (ppm))	0529	580.26	0.25000	587.82	57.289	20.061	88	-0.13	025.31	0
025	Iron (mg / kg (ppm))	0033	580.50	65.000	587.82	57.289	20.061	88	-0.13	025.43	0
025	Iron (mg / kg (ppm))	0511	582.00	4.0000	587.82	57.289	20.061	88	-0.10	025.41	0
025	Iron (mg / kg (ppm))	0208	582.35	24.100	587.82	57.289	20.061	88	-0.10	025.41	0
025	Iron (mg / kg (ppm))	0413	585.00	10.000	587.82	57.289	20.061	88	-0.05	025.42	0
025	Iron (mg / kg (ppm))	2068	592.57	9.4770	587.82	57.289	20.061	88	0.08	025.31	0
025	Iron (mg / kg (ppm))	0353	593.50	11.000	587.82	57.289	20.061	88	0.10	025.43	0
025	Iron (mg / kg (ppm))	0910	595.50	63.000	587.82	57.289	20.061	88	0.13	025.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
025	Iron (mg / kg (ppm))	0693	595.60	8.9471	587.82	57.289	20.061	88	0.14	025.42	0
025	Iron (mg / kg (ppm))	0675	595.94	6.7300	587.82	57.289	20.061	88	0.14	025.43	0
025	Iron (mg / kg (ppm))	0037	597.70	1.6000	587.82	57.289	20.061	88	0.17	025.42	0
025	Iron (mg / kg (ppm))	0563	600.91	16.199	587.82	57.289	20.061	88	0.23	025.31	0
025	Iron (mg / kg (ppm))	0083	601.50	15.000	587.82	57.289	20.061	88	0.24	025.43	0
025	Iron (mg / kg (ppm))	0918	602.42	12.430	587.82	57.289	20.061	88	0.25	025.43	0
025	Iron (mg / kg (ppm))	2012	603.43	1.4100	587.82	57.289	20.061	88	0.27	025.41	0
025	Iron (mg / kg (ppm))	0870	603.70	18.000	587.82	57.289	20.061	88	0.28	025.42	0
025	Iron (mg / kg (ppm))	0019	603.90	11.600	587.82	57.289	20.061	88	0.28	025.41	0
025	Iron (mg / kg (ppm))	0849	604.53	1.2500	587.82	57.289	20.061	88	0.29	025.31	0
025	Iron (mg / kg (ppm))	2004	606.50	15.000	587.82	57.289	20.061	88	0.33	025.41	0
025	Iron (mg / kg (ppm))	0505	607.50	35.000	587.82	57.289	20.061	88	0.34	025.31	0
025	Iron (mg / kg (ppm))	0148	610.50	13.000	587.82	57.289	20.061	88	0.40	025.41	0
025	Iron (mg / kg (ppm))	0553	613.00	16.000	587.82	57.289	20.061	88	0.44	025.53	0
025	Iron (mg / kg (ppm))	0964	616.45	1.1000	587.82	57.289	20.061	88	0.50	025.43	0
025	Iron (mg / kg (ppm))	0164	618.50	15.000	587.82	57.289	20.061	88	0.54	025.41	0
025	Iron (mg / kg (ppm))	0610	618.50	17.000	587.82	57.289	20.061	88	0.54	025.43	0
025	Iron (mg / kg (ppm))	0074	619.00	56.000	587.82	57.289	20.061	88	0.54	025.41	0
025	Iron (mg / kg (ppm))	0100	620.00	2.0000	587.82	57.289	20.061	88	0.56	025.99	0
025	Iron (mg / kg (ppm))	2146	621.05	11.700	587.82	57.289	20.061	88	0.58	025.51	0
025	Iron (mg / kg (ppm))	0098	621.20	43.400	587.82	57.289	20.061	88	0.58	025.41	0
025	Iron (mg / kg (ppm))	0520	622.00	12.000	587.82	57.289	20.061	88	0.60	025.41	0
025	Iron (mg / kg (ppm))	0366	623.50	19.000	587.82	57.289	20.061	88	0.62	025.42	0
025	Iron (mg / kg (ppm))	0504	624.00	37.200	587.82	57.289	20.061	88	0.63	025.33	0
025	Iron (mg / kg (ppm))	0098	624.00	37.400	587.82	57.289	20.061	88	0.63	025.43	0
025	Iron (mg / kg (ppm))	0297	625.00	14.000	587.82	57.289	20.061	88	0.65	025.43	0
025	Iron (mg / kg (ppm))	0874	626.50	5.0000	587.82	57.289	20.061	88	0.68	025.31	0
025	Iron (mg / kg (ppm))	0003	627.00	42.000	587.82	57.289	20.061	88	0.68	025.41	0
025	Iron (mg / kg (ppm))	0407	628.75	28.100	587.82	57.289	20.061	88	0.71	025.41	0
025	Iron (mg / kg (ppm))	0014	630.50	77.000	587.82	57.289	20.061	88	0.74	025.31	0
025	Iron (mg / kg (ppm))	0045	630.50	25.000	587.82	57.289	20.061	88	0.74	025.42	0
025	Iron (mg / kg (ppm))	0861	637.00	26.000	587.82	57.289	20.061	88	0.86	025.43	0
025	Iron (mg / kg (ppm))	0175	642.00	8.0000	587.82	57.289	20.061	88	0.95	025.31	0
025	Iron (mg / kg (ppm))	0354	642.02	32.550	587.82	57.289	20.061	88	0.95	025.31	0
025	Iron (mg / kg (ppm))	0948	642.47	30.826	587.82	57.289	20.061	88	0.95	025.33	0
025	Iron (mg / kg (ppm))	0035	642.50	11.000	587.82	57.289	20.061	88	0.95	025.42	0
025	Iron (mg / kg (ppm))	0692	677.00	40.000	587.82	57.289	20.061	88	1.56	025.42	0
025	Iron (mg / kg (ppm))	2053	680.75	16.500	587.82	57.289	20.061	88	1.62	025.42	0
025	Iron (mg / kg (ppm))	0504	681.40	6.4000	587.82	57.289	20.061	88	1.63	025.32	0
025	Iron (mg / kg (ppm))	0226	697.50	47.000	587.82	57.289	20.061	88	1.91	025.41	0
025	Iron (mg / kg (ppm))	0208	698.00	78.000	587.82	57.289	20.061	88	1.92	025.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
025	Iron (mg / kg (ppm))	0358	699.28	57.760	587.82	57.289	20.061	88	1.95	025.41	0
025	Iron (mg / kg (ppm))	0042	715.00	28.000	587.82	57.289	20.061	88	2.22	025.42	0
025	Iron (mg / kg (ppm))	0171	739.00	24.000	587.82	57.289	20.061	88	2.64	025.41	0
025	Iron (mg / kg (ppm))	0560	532.10	117.60	587.82	57.289	20.061	88	-0.97	025.42	1
025	Iron (mg / kg (ppm))	0049	663.29	133.60	587.82	57.289	20.061	88	1.32	025.41	1
025	Iron (mg / kg (ppm))	0169	0.00000	0.00000	587.82	57.289	20.061	88		025.43	4
027	Magnesium (%)	0026	0.16000	0.00000	0.23700	0.01293	0.00590	96	-5.96	027.42	0
027	Magnesium (%)	0689	0.20500	0.01000	0.23700	0.01293	0.00590	96	-2.48	027.31	0
027	Magnesium (%)	0504	0.20800	0.00000	0.23700	0.01293	0.00590	96	-2.24	027.33	0
027	Magnesium (%)	0028	0.21000	0.00000	0.23700	0.01293	0.00590	96	-2.09	027.43	0
027	Magnesium (%)	0202	0.21000	0.00000	0.23700	0.01293	0.00590	96	-2.09	027.43	0
027	Magnesium (%)	0001	0.21115	0.00150	0.23700	0.01293	0.00590	96	-2.00	027.31	0
027	Magnesium (%)	0939	0.21500	0.01000	0.23700	0.01293	0.00590	96	-1.70	027.31	0
027	Magnesium (%)	0910	0.21500	0.01000	0.23700	0.01293	0.00590	96	-1.70	027.41	0
027	Magnesium (%)	0345	0.21950	0.00900	0.23700	0.01293	0.00590	96	-1.35	027.43	0
027	Magnesium (%)	0510	0.22000	0.00000	0.23700	0.01293	0.00590	96	-1.32	027.43	0
027	Magnesium (%)	0590	0.22000	0.00000	0.23700	0.01293	0.00590	96	-1.32	027.99	0
027	Magnesium (%)	2129	0.22250	0.00500	0.23700	0.01293	0.00590	96	-1.12	027.42	0
027	Magnesium (%)	0563	0.22310	0.00320	0.23700	0.01293	0.00590	96	-1.08	027.31	0
027	Magnesium (%)	2055	0.22425	0.01290	0.23700	0.01293	0.00590	96	-0.99	027.53	0
027	Magnesium (%)	0148	0.22500	0.01000	0.23700	0.01293	0.00590	96	-0.93	027.41	0
027	Magnesium (%)	0265	0.22500	0.01000	0.23700	0.01293	0.00590	96	-0.93	027.42	0
027	Magnesium (%)	0014	0.22700	0.00200	0.23700	0.01293	0.00590	96	-0.77	027.31	0
027	Magnesium (%)	0918	0.22800	0.00600	0.23700	0.01293	0.00590	96	-0.70	027.43	0
027	Magnesium (%)	0154	0.22925	0.00130	0.23700	0.01293	0.00590	96	-0.60	027.52	0
027	Magnesium (%)	0142	0.22930	0.00360	0.23700	0.01293	0.00590	96	-0.60	027.31	0
027	Magnesium (%)	0629	0.22950	0.00100	0.23700	0.01293	0.00590	96	-0.58	027.43	0
027	Magnesium (%)	0263	0.22985	0.00270	0.23700	0.01293	0.00590	96	-0.55	027.32	0
027	Magnesium (%)	0505	0.23000	0.01000	0.23700	0.01293	0.00590	96	-0.54	027.31	0
027	Magnesium (%)	0013	0.23000	0.02000	0.23700	0.01293	0.00590	96	-0.54	027.43	0
027	Magnesium (%)	0629	0.23000	0.00000	0.23700	0.01293	0.00590	96	-0.54	027.31	0
027	Magnesium (%)	0049	0.23000	0.00000	0.23700	0.01293	0.00590	96	-0.54	027.41	0
027	Magnesium (%)	0511	0.23000	0.00000	0.23700	0.01293	0.00590	96	-0.54	027.41	0
027	Magnesium (%)	2012	0.23000	0.00000	0.23700	0.01293	0.00590	96	-0.54	027.41	0
027	Magnesium (%)	0685	0.23000	0.00000	0.23700	0.01293	0.00590	96	-0.54	027.42	0
027	Magnesium (%)	0035	0.23000	0.00000	0.23700	0.01293	0.00590	96	-0.54	027.43	0
027	Magnesium (%)	0870	0.23015	0.00030	0.23700	0.01293	0.00590	96	-0.53	027.42	0
027	Magnesium (%)	0964	0.23050	0.00500	0.23700	0.01293	0.00590	96	-0.50	027.43	0
027	Magnesium (%)	0043	0.23100	0.00200	0.23700	0.01293	0.00590	96	-0.46	027.43	0
027	Magnesium (%)	0955	0.23200	0.00800	0.23700	0.01293	0.00590	96	-0.39	027.44	0
027	Magnesium (%)	0047	0.23200	0.01200	0.23700	0.01293	0.00590	96	-0.39	027.52	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
027	Magnesium (%)	0504	0.23200	0.00000	0.23700	0.01293	0.00590	96	-0.39	027.32	0
027	Magnesium (%)	0560	0.23205	0.00250	0.23700	0.01293	0.00590	96	-0.38	027.42	0
027	Magnesium (%)	0186	0.23250	0.00300	0.23700	0.01293	0.00590	96	-0.35	027.52	0
027	Magnesium (%)	0208	0.23350	0.01900	0.23700	0.01293	0.00590	96	-0.27	027.31	0
027	Magnesium (%)	0164	0.23350	0.00500	0.23700	0.01293	0.00590	96	-0.27	027.41	0
027	Magnesium (%)	0045	0.23400	0.00400	0.23700	0.01293	0.00590	96	-0.23	027.42	0
027	Magnesium (%)	0175	0.23500	0.01000	0.23700	0.01293	0.00590	96	-0.15	027.31	0
027	Magnesium (%)	0354	0.23500	0.01000	0.23700	0.01293	0.00590	96	-0.15	027.31	0
027	Magnesium (%)	0098	0.23500	0.01000	0.23700	0.01293	0.00590	96	-0.15	027.41	0
027	Magnesium (%)	0878	0.23500	0.01000	0.23700	0.01293	0.00590	96	-0.15	027.41	0
027	Magnesium (%)	0098	0.23500	0.01000	0.23700	0.01293	0.00590	96	-0.15	027.43	0
027	Magnesium (%)	0425	0.23500	0.01000	0.23700	0.01293	0.00590	96	-0.15	027.43	0
027	Magnesium (%)	0861	0.23550	0.00500	0.23700	0.01293	0.00590	96	-0.12	027.43	0
027	Magnesium (%)	0948	0.23600	0.00400	0.23700	0.01293	0.00590	96	-0.08	027.33	0
027	Magnesium (%)	0139	0.23600	0.00140	0.23700	0.01293	0.00590	96	-0.08	027.41	0
027	Magnesium (%)	0009	0.23630	0.00160	0.23700	0.01293	0.00590	96	-0.05	027.42	0
027	Magnesium (%)	0650	0.23640	0.01420	0.23700	0.01293	0.00590	96	-0.05	027.31	0
027	Magnesium (%)	0572	0.23650	0.00500	0.23700	0.01293	0.00590	96	-0.04	027.53	0
027	Magnesium (%)	2004	0.23750	0.00100	0.23700	0.01293	0.00590	96	0.04	027.44	0
027	Magnesium (%)	0407	0.23855	0.00610	0.23700	0.01293	0.00590	96	0.12	027.41	0
027	Magnesium (%)	0034	0.23855	0.00770	0.23700	0.01293	0.00590	96	0.12	027.43	0
027	Magnesium (%)	2022	0.23900	0.00200	0.23700	0.01293	0.00590	96	0.15	027.31	0
027	Magnesium (%)	0208	0.23950	0.00900	0.23700	0.01293	0.00590	96	0.19	027.41	0
027	Magnesium (%)	0874	0.24000	0.00600	0.23700	0.01293	0.00590	96	0.23	027.31	0
027	Magnesium (%)	0035	0.24000	0.00000	0.23700	0.01293	0.00590	96	0.23	027.32	0
027	Magnesium (%)	0035	0.24000	0.00000	0.23700	0.01293	0.00590	96	0.23	027.33	0
027	Magnesium (%)	0074	0.24000	0.00000	0.23700	0.01293	0.00590	96	0.23	027.41	0
027	Magnesium (%)	0413	0.24000	0.02000	0.23700	0.01293	0.00590	96	0.23	027.41	0
027	Magnesium (%)	0035	0.24000	0.00000	0.23700	0.01293	0.00590	96	0.23	027.42	0
027	Magnesium (%)	0294	0.24000	0.00000	0.23700	0.01293	0.00590	96	0.23	027.42	0
027	Magnesium (%)	0610	0.24000	0.00000	0.23700	0.01293	0.00590	96	0.23	027.43	0
027	Magnesium (%)	0675	0.24000	0.02000	0.23700	0.01293	0.00590	96	0.23	027.43	0
027	Magnesium (%)	2146	0.24150	0.00300	0.23700	0.01293	0.00590	96	0.35	027.51	0
027	Magnesium (%)	0019	0.24500	0.01000	0.23700	0.01293	0.00590	96	0.62	027.41	0
027	Magnesium (%)	0190	0.24500	0.01000	0.23700	0.01293	0.00590	96	0.62	027.42	0
027	Magnesium (%)	0278	0.24500	0.01000	0.23700	0.01293	0.00590	96	0.62	027.42	0
027	Magnesium (%)	0598	0.24640	0.00620	0.23700	0.01293	0.00590	96	0.73	027.41	0
027	Magnesium (%)	0726	0.24700	0.00060	0.23700	0.01293	0.00590	96	0.77	027.42	0
027	Magnesium (%)	0186	0.24800	0.00600	0.23700	0.01293	0.00590	96	0.85	027.42	0
027	Magnesium (%)	0693	0.24925	0.00610	0.23700	0.01293	0.00590	96	0.95	027.42	0
027	Magnesium (%)	0169	0.25000	0.00000	0.23700	0.01293	0.00590	96	1.01	027.32	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
027	Magnesium (%)	0003	0.25000	0.00000	0.23700	0.01293	0.00590	96	1.01	027.41	0
027	Magnesium (%)	0520	0.25000	0.00000	0.23700	0.01293	0.00590	96	1.01	027.41	0
027	Magnesium (%)	0037	0.25000	0.02000	0.23700	0.01293	0.00590	96	1.01	027.42	0
027	Magnesium (%)	0366	0.25000	0.00000	0.23700	0.01293	0.00590	96	1.01	027.42	0
027	Magnesium (%)	0353	0.25000	0.02000	0.23700	0.01293	0.00590	96	1.01	027.43	0
027	Magnesium (%)	0682	0.25000	0.00000	0.23700	0.01293	0.00590	96	1.01	027.43	0
027	Magnesium (%)	0100	0.25000	0.00000	0.23700	0.01293	0.00590	96	1.01	027.99	0
027	Magnesium (%)	0357	0.25245	0.00570	0.23700	0.01293	0.00590	96	1.20	027.42	0
027	Magnesium (%)	0006	0.25300	0.00000	0.23700	0.01293	0.00590	96	1.24	027.43	0
027	Magnesium (%)	0619	0.25350	0.01100	0.23700	0.01293	0.00590	96	1.28	027.41	0
027	Magnesium (%)	0358	0.25500	0.01000	0.23700	0.01293	0.00590	96	1.39	027.41	0
027	Magnesium (%)	0083	0.25500	0.01000	0.23700	0.01293	0.00590	96	1.39	027.43	0
027	Magnesium (%)	2089	0.25500	0.01000	0.23700	0.01293	0.00590	96	1.39	027.43	0
027	Magnesium (%)	0553	0.25650	0.01900	0.23700	0.01293	0.00590	96	1.51	027.53	0
027	Magnesium (%)	0187	0.26440	0.00140	0.23700	0.01293	0.00590	96	2.12	027.42	0
027	Magnesium (%)	0226	0.26500	0.01000	0.23700	0.01293	0.00590	96	2.17	027.41	0
027	Magnesium (%)	0692	0.27500	0.01000	0.23700	0.01293	0.00590	96	2.94	027.42	0
027	Magnesium (%)	0968	0.28200	0.00200	0.23700	0.01293	0.00590	96	3.48	027.43	0
027	Magnesium (%)	2053	0.30500	0.03000	0.23700	0.01293	0.00590	96	5.26	027.42	0
027	Magnesium (%)	0405	0.40850	0.00300	0.23700	0.01293	0.00590	96	13.27	027.41	0
027	Magnesium (%)	0297	0.23000	0.04000	0.23700	0.01293	0.00590	96	-0.54	027.43	1
027	Magnesium (%)	0529	0.25000	0.04000	0.23700	0.01293	0.00590	96	1.01	027.31	1
027	Magnesium (%)	0042	0.25400	0.04600	0.23700	0.01293	0.00590	96	1.32	027.42	1
027	Magnesium (%)	0536	0.64000	0.02000	0.23700	0.01293	0.00590	96	31.18	027.31	2
027	Magnesium (%)	0171	0.71400	0.00400	0.23700	0.01293	0.00590	96	36.90	027.41	2
028	Manganese (mg / kg (ppm))	0026	129.50	29.000	178.52	12.748	7.0021	93	-3.85	028.42	0
028	Manganese (mg / kg (ppm))	2145	132.00	0.00000	178.52	12.748	7.0021	93	-3.65	028.41	0
028	Manganese (mg / kg (ppm))	0955	154.50	3.0000	178.52	12.748	7.0021	93	-1.88	028.44	0
028	Manganese (mg / kg (ppm))	0685	157.18	0.60000	178.52	12.748	7.0021	93	-1.67	028.42	0
028	Manganese (mg / kg (ppm))	0689	158.95	4.9000	178.52	12.748	7.0021	93	-1.54	028.31	0
028	Manganese (mg / kg (ppm))	0619	159.50	3.0000	178.52	12.748	7.0021	93	-1.49	028.41	0
028	Manganese (mg / kg (ppm))	0510	159.50	5.0000	178.52	12.748	7.0021	93	-1.49	028.43	0
028	Manganese (mg / kg (ppm))	2022	161.00	6.0000	178.52	12.748	7.0021	93	-1.37	028.31	0
028	Manganese (mg / kg (ppm))	0265	162.50	9.0000	178.52	12.748	7.0021	93	-1.26	028.42	0
028	Manganese (mg / kg (ppm))	0043	164.00	6.0000	178.52	12.748	7.0021	93	-1.14	028.43	0
028	Manganese (mg / kg (ppm))	0529	164.82	12.130	178.52	12.748	7.0021	93	-1.08	028.31	0
028	Manganese (mg / kg (ppm))	0629	165.00	0.00000	178.52	12.748	7.0021	93	-1.06	028.43	0
028	Manganese (mg / kg (ppm))	0034	165.37	12.220	178.52	12.748	7.0021	93	-1.03	028.43	0
028	Manganese (mg / kg (ppm))	0921	166.20	0.00000	178.52	12.748	7.0021	93	-0.97	028.31	0
028	Manganese (mg / kg (ppm))	0148	167.00	4.0000	178.52	12.748	7.0021	93	-0.90	028.41	0
028	Manganese (mg / kg (ppm))	0278	167.50	5.0000	178.52	12.748	7.0021	93	-0.86	028.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
028	Manganese (mg / kg (ppm))	0035	167.50	17.000	178.52	12.748	7.0021	93	-0.86	028.99	0
028	Manganese (mg / kg (ppm))	2055	168.45	26.100	178.52	12.748	7.0021	93	-0.79	028.53	0
028	Manganese (mg / kg (ppm))	0511	168.50	3.0000	178.52	12.748	7.0021	93	-0.79	028.41	0
028	Manganese (mg / kg (ppm))	0520	169.50	1.0000	178.52	12.748	7.0021	93	-0.71	028.41	0
028	Manganese (mg / kg (ppm))	0910	169.50	5.0000	178.52	12.748	7.0021	93	-0.71	028.41	0
028	Manganese (mg / kg (ppm))	0505	170.50	4.0000	178.52	12.748	7.0021	93	-0.63	028.31	0
028	Manganese (mg / kg (ppm))	0358	170.74	1.0600	178.52	12.748	7.0021	93	-0.61	028.41	0
028	Manganese (mg / kg (ppm))	0878	171.00	2.0000	178.52	12.748	7.0021	93	-0.59	028.41	0
028	Manganese (mg / kg (ppm))	0563	171.37	1.0205	178.52	12.748	7.0021	93	-0.56	028.31	0
028	Manganese (mg / kg (ppm))	0629	172.00	4.0000	178.52	12.748	7.0021	93	-0.51	028.31	0
028	Manganese (mg / kg (ppm))	0590	172.00	0.00000	178.52	12.748	7.0021	93	-0.51	028.99	0
028	Manganese (mg / kg (ppm))	0948	172.10	3.3540	178.52	12.748	7.0021	93	-0.50	028.33	0
028	Manganese (mg / kg (ppm))	0504	172.15	12.500	178.52	12.748	7.0021	93	-0.50	028.33	0
028	Manganese (mg / kg (ppm))	0297	172.50	25.000	178.52	12.748	7.0021	93	-0.47	028.43	0
028	Manganese (mg / kg (ppm))	0098	172.65	15.500	178.52	12.748	7.0021	93	-0.46	028.41	0
028	Manganese (mg / kg (ppm))	0918	172.69	6.6900	178.52	12.748	7.0021	93	-0.46	028.43	0
028	Manganese (mg / kg (ppm))	0413	173.00	2.0000	178.52	12.748	7.0021	93	-0.43	028.42	0
028	Manganese (mg / kg (ppm))	0186	173.00	4.0000	178.52	12.748	7.0021	93	-0.43	028.52	0
028	Manganese (mg / kg (ppm))	2012	173.31	0.85000	178.52	12.748	7.0021	93	-0.41	028.41	0
028	Manganese (mg / kg (ppm))	0175	174.00	4.0000	178.52	12.748	7.0021	93	-0.35	028.31	0
028	Manganese (mg / kg (ppm))	0964	174.40	15.600	178.52	12.748	7.0021	93	-0.32	028.43	0
028	Manganese (mg / kg (ppm))	0049	174.52	26.110	178.52	12.748	7.0021	93	-0.31	028.41	0
028	Manganese (mg / kg (ppm))	0693	175.00	5.5862	178.52	12.748	7.0021	93	-0.28	028.42	0
028	Manganese (mg / kg (ppm))	0572	175.50	7.0000	178.52	12.748	7.0021	93	-0.24	028.53	0
028	Manganese (mg / kg (ppm))	0425	175.65	0.90000	178.52	12.748	7.0021	93	-0.23	028.43	0
028	Manganese (mg / kg (ppm))	0939	175.68	2.5300	178.52	12.748	7.0021	93	-0.22	028.31	0
028	Manganese (mg / kg (ppm))	0208	176.00	4.6000	178.52	12.748	7.0021	93	-0.20	028.41	0
028	Manganese (mg / kg (ppm))	0001	177.05	14.300	178.52	12.748	7.0021	93	-0.12	028.31	0
028	Manganese (mg / kg (ppm))	0405	177.26	0.11000	178.52	12.748	7.0021	93	-0.10	028.41	0
028	Manganese (mg / kg (ppm))	0968	177.50	1.0000	178.52	12.748	7.0021	93	-0.08	028.43	0
028	Manganese (mg / kg (ppm))	2004	177.50	5.0000	178.52	12.748	7.0021	93	-0.08	028.44	0
028	Manganese (mg / kg (ppm))	0164	178.00	4.0000	178.52	12.748	7.0021	93	-0.04	028.41	0
028	Manganese (mg / kg (ppm))	0074	178.50	1.0000	178.52	12.748	7.0021	93	0.00	028.41	0
028	Manganese (mg / kg (ppm))	0345	179.00	2.0000	178.52	12.748	7.0021	93	0.04	028.43	0
028	Manganese (mg / kg (ppm))	2089	179.26	4.4400	178.52	12.748	7.0021	93	0.06	028.43	0
028	Manganese (mg / kg (ppm))	0208	179.50	1.0000	178.52	12.748	7.0021	93	0.08	028.31	0
028	Manganese (mg / kg (ppm))	0013	180.00	12.000	178.52	12.748	7.0021	93	0.12	028.43	0
028	Manganese (mg / kg (ppm))	0553	180.00	10.000	178.52	12.748	7.0021	93	0.12	028.53	0
028	Manganese (mg / kg (ppm))	0045	180.50	9.0000	178.52	12.748	7.0021	93	0.16	028.42	0
028	Manganese (mg / kg (ppm))	0098	180.75	12.300	178.52	12.748	7.0021	93	0.17	028.43	0
028	Manganese (mg / kg (ppm))	0035	181.00	8.0000	178.52	12.748	7.0021	93	0.19	028.32	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
028	Manganese (mg / kg (ppm))	0560	181.10	6.2000	178.52	12.748	7.0021	93	0.20	028.42	0
028	Manganese (mg / kg (ppm))	0861	181.50	3.0000	178.52	12.748	7.0021	93	0.23	028.43	0
028	Manganese (mg / kg (ppm))	0354	182.69	0.65000	178.52	12.748	7.0021	93	0.33	028.31	0
028	Manganese (mg / kg (ppm))	0407	182.75	0.51390	178.52	12.748	7.0021	93	0.33	028.41	0
028	Manganese (mg / kg (ppm))	0035	183.00	6.0000	178.52	12.748	7.0021	93	0.35	028.42	0
028	Manganese (mg / kg (ppm))	0083	183.00	4.0000	178.52	12.748	7.0021	93	0.35	028.43	0
028	Manganese (mg / kg (ppm))	0100	183.00	6.0000	178.52	12.748	7.0021	93	0.35	028.99	0
028	Manganese (mg / kg (ppm))	0003	183.50	7.0000	178.52	12.748	7.0021	93	0.39	028.41	0
028	Manganese (mg / kg (ppm))	2068	184.03	0.17900	178.52	12.748	7.0021	93	0.43	028.31	0
028	Manganese (mg / kg (ppm))	0682	184.64	0.00000	178.52	12.748	7.0021	93	0.48	028.43	0
028	Manganese (mg / kg (ppm))	0019	184.90	8.6000	178.52	12.748	7.0021	93	0.50	028.41	0
028	Manganese (mg / kg (ppm))	0037	185.45	23.900	178.52	12.748	7.0021	93	0.54	028.42	0
028	Manganese (mg / kg (ppm))	0598	186.05	10.300	178.52	12.748	7.0021	93	0.59	028.41	0
028	Manganese (mg / kg (ppm))	0870	186.20	3.2000	178.52	12.748	7.0021	93	0.60	028.42	0
028	Manganese (mg / kg (ppm))	0187	187.23	1.7400	178.52	12.748	7.0021	93	0.68	028.42	0
028	Manganese (mg / kg (ppm))	0353	188.50	7.0000	178.52	12.748	7.0021	93	0.78	028.43	0
028	Manganese (mg / kg (ppm))	0675	188.91	6.0400	178.52	12.748	7.0021	93	0.81	028.43	0
028	Manganese (mg / kg (ppm))	0226	189.00	4.0000	178.52	12.748	7.0021	93	0.82	028.41	0
028	Manganese (mg / kg (ppm))	0035	189.00	4.0000	178.52	12.748	7.0021	93	0.82	028.43	0
028	Manganese (mg / kg (ppm))	0610	189.50	17.000	178.52	12.748	7.0021	93	0.86	028.43	0
028	Manganese (mg / kg (ppm))	0027	189.88	4.8210	178.52	12.748	7.0021	93	0.89	028.43	0
028	Manganese (mg / kg (ppm))	2053	193.45	21.300	178.52	12.748	7.0021	93	1.17	028.42	0
028	Manganese (mg / kg (ppm))	0357	193.71	3.1110	178.52	12.748	7.0021	93	1.19	028.42	0
028	Manganese (mg / kg (ppm))	0726	194.80	0.20000	178.52	12.748	7.0021	93	1.28	028.42	0
028	Manganese (mg / kg (ppm))	0202	195.55	15.940	178.52	12.748	7.0021	93	1.34	028.43	0
028	Manganese (mg / kg (ppm))	0190	196.75	3.3000	178.52	12.748	7.0021	93	1.43	028.42	0
028	Manganese (mg / kg (ppm))	0504	197.50	6.4000	178.52	12.748	7.0021	93	1.49	028.32	0
028	Manganese (mg / kg (ppm))	0366	197.50	5.0000	178.52	12.748	7.0021	93	1.49	028.42	0
028	Manganese (mg / kg (ppm))	0047	203.13	0.34000	178.52	12.748	7.0021	93	1.93	028.52	0
028	Manganese (mg / kg (ppm))	0294	204.18	10.360	178.52	12.748	7.0021	93	2.01	028.42	0
028	Manganese (mg / kg (ppm))	0042	208.00	34.000	178.52	12.748	7.0021	93	2.31	028.42	0
028	Manganese (mg / kg (ppm))	0186	211.00	18.000	178.52	12.748	7.0021	93	2.55	028.42	0
028	Manganese (mg / kg (ppm))	0692	216.50	7.0000	178.52	12.748	7.0021	93	2.98	028.42	0
028	Manganese (mg / kg (ppm))	2146	219.80	6.0000	178.52	12.748	7.0021	93	3.24	028.51	0
028	Manganese (mg / kg (ppm))	0536	405.75	6.7000	178.52	12.748	7.0021	93	17.82	028.31	0
028	Manganese (mg / kg (ppm))	0171	407.50	5.0000	178.52	12.748	7.0021	93	17.96	028.41	0
028	Manganese (mg / kg (ppm))	0009	1,611.4	74.749	178.52	12.748	7.0021	93	112.40	028.42	2
028	Manganese (mg / kg (ppm))	0169	0.00000	0.00000	178.52	12.748	7.0021	93		028.43	4
031	Phosphorus (%)	2022	0.53500	0.01000	0.71895	0.03781	0.01813	133	-4.86	031.01	0
031	Phosphorus (%)	0536	0.55000	0.06000	0.71895	0.03781	0.01813	133	-4.47	031.06	0
031	Phosphorus (%)	0171	0.59750	0.00500	0.71895	0.03781	0.01813	133	-3.21	031.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	2055	0.62215	0.04190	0.71895	0.03781	0.01813	133	-2.56	031.53	0
031	Phosphorus (%)	0026	0.63000	0.10000	0.71895	0.03781	0.01813	133	-2.35	031.42	0
031	Phosphorus (%)	0028	0.63500	0.01000	0.71895	0.03781	0.01813	133	-2.22	031.43	0
031	Phosphorus (%)	2129	0.64600	0.00400	0.71895	0.03781	0.01813	133	-1.93	031.42	0
031	Phosphorus (%)	0233	0.65000	0.04000	0.71895	0.03781	0.01813	133	-1.82	031.01	0
031	Phosphorus (%)	0187	0.65370	0.00040	0.71895	0.03781	0.01813	133	-1.73	031.42	0
031	Phosphorus (%)	0043	0.65500	0.01000	0.71895	0.03781	0.01813	133	-1.69	031.43	0
031	Phosphorus (%)	0108	0.66000	0.06000	0.71895	0.03781	0.01813	133	-1.56	031.01	0
031	Phosphorus (%)	0511	0.66500	0.09000	0.71895	0.03781	0.01813	133	-1.43	031.01	0
031	Phosphorus (%)	0940	0.66500	0.01000	0.71895	0.03781	0.01813	133	-1.43	031.01	0
031	Phosphorus (%)	0610	0.66500	0.01000	0.71895	0.03781	0.01813	133	-1.43	031.43	0
031	Phosphorus (%)	0152	0.67000	0.04000	0.71895	0.03781	0.01813	133	-1.29	031.01	0
031	Phosphorus (%)	0278	0.67000	0.06000	0.71895	0.03781	0.01813	133	-1.29	031.42	0
031	Phosphorus (%)	0910	0.67000	0.02000	0.71895	0.03781	0.01813	133	-1.29	031.41	0
031	Phosphorus (%)	0870	0.67460	0.02320	0.71895	0.03781	0.01813	133	-1.17	031.42	0
031	Phosphorus (%)	0035	0.67500	0.01000	0.71895	0.03781	0.01813	133	-1.16	031.01	0
031	Phosphorus (%)	0689	0.67500	0.01000	0.71895	0.03781	0.01813	133	-1.16	031.01	0
031	Phosphorus (%)	0297	0.67500	0.07000	0.71895	0.03781	0.01813	133	-1.16	031.43	0
031	Phosphorus (%)	0874	0.67745	0.00210	0.71895	0.03781	0.01813	133	-1.10	031.01	0
031	Phosphorus (%)	0186	0.67750	0.01100	0.71895	0.03781	0.01813	133	-1.10	031.52	0
031	Phosphorus (%)	0100	0.68000	0.02000	0.71895	0.03781	0.01813	133	-1.03	031.99	0
031	Phosphorus (%)	2145	0.68000	0.00000	0.71895	0.03781	0.01813	133	-1.03	031.41	0
031	Phosphorus (%)	0560	0.68345	0.00750	0.71895	0.03781	0.01813	133	-0.94	031.42	0
031	Phosphorus (%)	0034	0.68465	0.02830	0.71895	0.03781	0.01813	133	-0.91	031.43	0
031	Phosphorus (%)	0265	0.69000	0.00000	0.71895	0.03781	0.01813	133	-0.77	031.42	0
031	Phosphorus (%)	0294	0.69000	0.02000	0.71895	0.03781	0.01813	133	-0.77	031.42	0
031	Phosphorus (%)	0033	0.69000	0.04000	0.71895	0.03781	0.01813	133	-0.77	031.43	0
031	Phosphorus (%)	0202	0.69000	0.02000	0.71895	0.03781	0.01813	133	-0.77	031.43	0
031	Phosphorus (%)	0629	0.69000	0.00000	0.71895	0.03781	0.01813	133	-0.77	031.43	0
031	Phosphorus (%)	0897	0.69500	0.01000	0.71895	0.03781	0.01813	133	-0.63	031.01	0
031	Phosphorus (%)	0019	0.69500	0.01000	0.71895	0.03781	0.01813	133	-0.63	031.41	0
031	Phosphorus (%)	2089	0.69500	0.03000	0.71895	0.03781	0.01813	133	-0.63	031.43	0
031	Phosphorus (%)	0623	0.69750	0.05280	0.71895	0.03781	0.01813	133	-0.57	031.01	0
031	Phosphorus (%)	0144	0.70000	0.00000	0.71895	0.03781	0.01813	133	-0.50	031.41	0
031	Phosphorus (%)	0154	0.70010	0.00440	0.71895	0.03781	0.01813	133	-0.50	031.52	0
031	Phosphorus (%)	0142	0.70270	0.01060	0.71895	0.03781	0.01813	133	-0.43	031.01	0
031	Phosphorus (%)	0849	0.70500	0.01000	0.71895	0.03781	0.01813	133	-0.37	031.01	0
031	Phosphorus (%)	0848	0.70500	0.01000	0.71895	0.03781	0.01813	133	-0.37	031.41	0
031	Phosphorus (%)	0035	0.70500	0.01000	0.71895	0.03781	0.01813	133	-0.37	031.43	0
031	Phosphorus (%)	0620	0.70625	0.04870	0.71895	0.03781	0.01813	133	-0.34	031.01	0
031	Phosphorus (%)	0098	0.70670	0.02960	0.71895	0.03781	0.01813	133	-0.32	031.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	0345	0.70950	0.00500	0.71895	0.03781	0.01813	133	-0.25	031.43	0
031	Phosphorus (%)	0552	0.71000	0.00000	0.71895	0.03781	0.01813	133	-0.24	031.01	0
031	Phosphorus (%)	0939	0.71000	0.06000	0.71895	0.03781	0.01813	133	-0.24	031.01	0
031	Phosphorus (%)	0148	0.71000	0.02000	0.71895	0.03781	0.01813	133	-0.24	031.41	0
031	Phosphorus (%)	0045	0.71000	0.01600	0.71895	0.03781	0.01813	133	-0.24	031.42	0
031	Phosphorus (%)	2152	0.71000	0.02000	0.71895	0.03781	0.01813	133	-0.24	031.44	0
031	Phosphorus (%)	2006	0.71050	0.00100	0.71895	0.03781	0.01813	133	-0.22	031.01	0
031	Phosphorus (%)	0186	0.71100	0.03000	0.71895	0.03781	0.01813	133	-0.21	031.42	0
031	Phosphorus (%)	0006	0.71150	0.00100	0.71895	0.03781	0.01813	133	-0.20	031.43	0
031	Phosphorus (%)	2009	0.71225	0.00050	0.71895	0.03781	0.01813	133	-0.18	031.01	0
031	Phosphorus (%)	0208	0.71300	0.00200	0.71895	0.03781	0.01813	133	-0.16	031.41	0
031	Phosphorus (%)	0968	0.71450	0.00700	0.71895	0.03781	0.01813	133	-0.12	031.43	0
031	Phosphorus (%)	0194	0.71500	0.01000	0.71895	0.03781	0.01813	133	-0.10	031.01	0
031	Phosphorus (%)	0626	0.71550	0.02500	0.71895	0.03781	0.01813	133	-0.09	031.01	0
031	Phosphorus (%)	0622	0.71600	0.00200	0.71895	0.03781	0.01813	133	-0.08	031.01	0
031	Phosphorus (%)	0027	0.71750	0.02900	0.71895	0.03781	0.01813	133	-0.04	031.43	0
031	Phosphorus (%)	0918	0.71750	0.01500	0.71895	0.03781	0.01813	133	-0.04	031.43	0
031	Phosphorus (%)	0693	0.71870	0.02340	0.71895	0.03781	0.01813	133	-0.01	031.42	0
031	Phosphorus (%)	0354	0.72000	0.00000	0.71895	0.03781	0.01813	133	0.03	031.01	0
031	Phosphorus (%)	0629	0.72000	0.00000	0.71895	0.03781	0.01813	133	0.03	031.01	0
031	Phosphorus (%)	0504	0.72000	0.00000	0.71895	0.03781	0.01813	133	0.03	031.03	0
031	Phosphorus (%)	0298	0.72000	0.02000	0.71895	0.03781	0.01813	133	0.03	031.41	0
031	Phosphorus (%)	2012	0.72000	0.00000	0.71895	0.03781	0.01813	133	0.03	031.41	0
031	Phosphorus (%)	0035	0.72000	0.02000	0.71895	0.03781	0.01813	133	0.03	031.42	0
031	Phosphorus (%)	0263	0.72110	0.00620	0.71895	0.03781	0.01813	133	0.06	031.01	0
031	Phosphorus (%)	0208	0.72200	0.04800	0.71895	0.03781	0.01813	133	0.08	031.01	0
031	Phosphorus (%)	0505	0.72500	0.01000	0.71895	0.03781	0.01813	133	0.16	031.02	0
031	Phosphorus (%)	0675	0.72500	0.01000	0.71895	0.03781	0.01813	133	0.16	031.43	0
031	Phosphorus (%)	0205	0.72750	0.00700	0.71895	0.03781	0.01813	133	0.23	031.42	0
031	Phosphorus (%)	0726	0.72850	0.00080	0.71895	0.03781	0.01813	133	0.25	031.42	0
031	Phosphorus (%)	0009	0.72910	0.00400	0.71895	0.03781	0.01813	133	0.27	031.42	0
031	Phosphorus (%)	0651	0.72950	0.01500	0.71895	0.03781	0.01813	133	0.28	031.01	0
031	Phosphorus (%)	0175	0.73000	0.00000	0.71895	0.03781	0.01813	133	0.29	031.01	0
031	Phosphorus (%)	0590	0.73000	0.02000	0.71895	0.03781	0.01813	133	0.29	031.01	0
031	Phosphorus (%)	0650	0.73000	0.00000	0.71895	0.03781	0.01813	133	0.29	031.01	0
031	Phosphorus (%)	0049	0.73000	0.00000	0.71895	0.03781	0.01813	133	0.29	031.41	0
031	Phosphorus (%)	0520	0.73000	0.02000	0.71895	0.03781	0.01813	133	0.29	031.41	0
031	Phosphorus (%)	0190	0.73000	0.02000	0.71895	0.03781	0.01813	133	0.29	031.42	0
031	Phosphorus (%)	0036	0.73080	0.00960	0.71895	0.03781	0.01813	133	0.31	031.03	0
031	Phosphorus (%)	0653	0.73100	0.00200	0.71895	0.03781	0.01813	133	0.32	031.01	0
031	Phosphorus (%)	2004	0.73150	0.02300	0.71895	0.03781	0.01813	133	0.33	031.44	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	0139	0.73300	0.00000	0.71895	0.03781	0.01813	133	0.37	031.41	0
031	Phosphorus (%)	0955	0.73350	0.01500	0.71895	0.03781	0.01813	133	0.38	031.44	0
031	Phosphorus (%)	0036	0.73380	0.00380	0.71895	0.03781	0.01813	133	0.39	031.43	0
031	Phosphorus (%)	2146	0.73400	0.02000	0.71895	0.03781	0.01813	133	0.40	031.51	0
031	Phosphorus (%)	0895	0.73500	0.01000	0.71895	0.03781	0.01813	133	0.42	031.00	0
031	Phosphorus (%)	0683	0.73500	0.03000	0.71895	0.03781	0.01813	133	0.42	031.01	0
031	Phosphorus (%)	0878	0.73500	0.01000	0.71895	0.03781	0.01813	133	0.42	031.01	0
031	Phosphorus (%)	0903	0.73500	0.01000	0.71895	0.03781	0.01813	133	0.42	031.01	0
031	Phosphorus (%)	2093	0.73500	0.01000	0.71895	0.03781	0.01813	133	0.42	031.01	0
031	Phosphorus (%)	0164	0.73500	0.01000	0.71895	0.03781	0.01813	133	0.42	031.41	0
031	Phosphorus (%)	2062	0.73625	0.00090	0.71895	0.03781	0.01813	133	0.46	031.01	0
031	Phosphorus (%)	0885	0.73660	0.01300	0.71895	0.03781	0.01813	133	0.47	031.01	0
031	Phosphorus (%)	0881	0.73950	0.00100	0.71895	0.03781	0.01813	133	0.54	031.01	0
031	Phosphorus (%)	2068	0.74000	0.00000	0.71895	0.03781	0.01813	133	0.56	031.01	0
031	Phosphorus (%)	0043	0.74000	0.00000	0.71895	0.03781	0.01813	133	0.56	031.02	0
031	Phosphorus (%)	2053	0.74000	0.02000	0.71895	0.03781	0.01813	133	0.56	031.02	0
031	Phosphorus (%)	0169	0.74000	0.00000	0.71895	0.03781	0.01813	133	0.56	031.03	0
031	Phosphorus (%)	0948	0.74450	0.00100	0.71895	0.03781	0.01813	133	0.68	031.01	0
031	Phosphorus (%)	0413	0.74500	0.01000	0.71895	0.03781	0.01813	133	0.69	031.41	0
031	Phosphorus (%)	0685	0.74500	0.01000	0.71895	0.03781	0.01813	133	0.69	031.41	0
031	Phosphorus (%)	0353	0.74500	0.01000	0.71895	0.03781	0.01813	133	0.69	031.43	0
031	Phosphorus (%)	0425	0.74500	0.01000	0.71895	0.03781	0.01813	133	0.69	031.43	0
031	Phosphorus (%)	0964	0.74750	0.01500	0.71895	0.03781	0.01813	133	0.76	031.43	0
031	Phosphorus (%)	0407	0.75100	0.03140	0.71895	0.03781	0.01813	133	0.85	031.41	0
031	Phosphorus (%)	0687	0.75500	0.07000	0.71895	0.03781	0.01813	133	0.95	031.01	0
031	Phosphorus (%)	0074	0.75500	0.01000	0.71895	0.03781	0.01813	133	0.95	031.41	0
031	Phosphorus (%)	0037	0.75500	0.01000	0.71895	0.03781	0.01813	133	0.95	031.42	0
031	Phosphorus (%)	0083	0.75500	0.01000	0.71895	0.03781	0.01813	133	0.95	031.43	0
031	Phosphorus (%)	0553	0.75550	0.00900	0.71895	0.03781	0.01813	133	0.97	031.53	0
031	Phosphorus (%)	0504	0.75750	0.04900	0.71895	0.03781	0.01813	133	1.02	031.42	0
031	Phosphorus (%)	0042	0.75800	0.19600	0.71895	0.03781	0.01813	133	1.03	031.42	0
031	Phosphorus (%)	0366	0.76000	0.02000	0.71895	0.03781	0.01813	133	1.09	031.42	0
031	Phosphorus (%)	0682	0.76000	0.00000	0.71895	0.03781	0.01813	133	1.09	031.43	0
031	Phosphorus (%)	0619	0.76250	0.02700	0.71895	0.03781	0.01813	133	1.15	031.41	0
031	Phosphorus (%)	0563	0.76355	0.00750	0.71895	0.03781	0.01813	133	1.18	031.01	0
031	Phosphorus (%)	0572	0.76650	0.03900	0.71895	0.03781	0.01813	133	1.26	031.53	0
031	Phosphorus (%)	0893	0.78000	0.00000	0.71895	0.03781	0.01813	133	1.61	031.00	0
031	Phosphorus (%)	0861	0.78150	0.01300	0.71895	0.03781	0.01813	133	1.65	031.43	0
031	Phosphorus (%)	2109	0.78450	0.00500	0.71895	0.03781	0.01813	133	1.73	031.01	0
031	Phosphorus (%)	0357	0.78585	0.00490	0.71895	0.03781	0.01813	133	1.77	031.42	0
031	Phosphorus (%)	0098	0.78830	0.06000	0.71895	0.03781	0.01813	133	1.83	031.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	0226	0.79000	0.00000	0.71895	0.03781	0.01813	133	1.88	031.41	0
031	Phosphorus (%)	0003	0.80000	0.02000	0.71895	0.03781	0.01813	133	2.14	031.41	0
031	Phosphorus (%)	0358	0.80000	0.02000	0.71895	0.03781	0.01813	133	2.14	031.41	0
031	Phosphorus (%)	0510	0.80000	0.00000	0.71895	0.03781	0.01813	133	2.14	031.43	0
031	Phosphorus (%)	0692	0.82500	0.01000	0.71895	0.03781	0.01813	133	2.80	031.42	0
031	Phosphorus (%)	0598	0.82530	0.01780	0.71895	0.03781	0.01813	133	2.81	031.41	0
031	Phosphorus (%)	0405	0.84500	0.01000	0.71895	0.03781	0.01813	133	3.33	031.41	0
031	Phosphorus (%)	0529	1.3650	1.2500	0.71895	0.03781	0.01813	133	17.09	031.01	2
032	Potassium (%)	0026	0.95500	0.01000	1.3647	0.06623	0.03011	101	-6.19	032.42	0
032	Potassium (%)	0874	1.2200	0.02000	1.3647	0.06623	0.03011	101	-2.19	032.31	0
032	Potassium (%)	0505	1.2300	0.02000	1.3647	0.06623	0.03011	101	-2.03	032.31	0
032	Potassium (%)	0629	1.2300	0.00000	1.3647	0.06623	0.03011	101	-2.03	032.43	0
032	Potassium (%)	0629	1.2350	0.03000	1.3647	0.06623	0.03011	101	-1.96	032.31	0
032	Potassium (%)	0035	1.2600	0.00000	1.3647	0.06623	0.03011	101	-1.58	032.43	0
032	Potassium (%)	0511	1.2650	0.01000	1.3647	0.06623	0.03011	101	-1.51	032.41	0
032	Potassium (%)	0028	1.2650	0.01000	1.3647	0.06623	0.03011	101	-1.51	032.43	0
032	Potassium (%)	0910	1.2700	0.06000	1.3647	0.06623	0.03011	101	-1.43	032.41	0
032	Potassium (%)	2022	1.2795	0.01300	1.3647	0.06623	0.03011	101	-1.29	032.31	0
032	Potassium (%)	0529	1.2800	0.14000	1.3647	0.06623	0.03011	101	-1.28	032.31	0
032	Potassium (%)	0013	1.2850	0.11000	1.3647	0.06623	0.03011	101	-1.20	032.43	0
032	Potassium (%)	0278	1.2900	0.06000	1.3647	0.06623	0.03011	101	-1.13	032.42	0
032	Potassium (%)	0964	1.2940	0.10000	1.3647	0.06623	0.03011	101	-1.07	032.43	0
032	Potassium (%)	0019	1.3000	0.06000	1.3647	0.06623	0.03011	101	-0.98	032.41	0
032	Potassium (%)	0202	1.3050	0.01000	1.3647	0.06623	0.03011	101	-0.90	032.43	0
032	Potassium (%)	0263	1.3087	0.01100	1.3647	0.06623	0.03011	101	-0.85	032.32	0
032	Potassium (%)	0098	1.3100	0.00000	1.3647	0.06623	0.03011	101	-0.83	032.43	0
032	Potassium (%)	0345	1.3100	0.04000	1.3647	0.06623	0.03011	101	-0.83	032.43	0
032	Potassium (%)	0861	1.3100	0.10000	1.3647	0.06623	0.03011	101	-0.83	032.43	0
032	Potassium (%)	0968	1.3170	0.00800	1.3647	0.06623	0.03011	101	-0.72	032.43	0
032	Potassium (%)	0939	1.3200	0.08000	1.3647	0.06623	0.03011	101	-0.68	032.31	0
032	Potassium (%)	0003	1.3200	0.02000	1.3647	0.06623	0.03011	101	-0.68	032.41	0
032	Potassium (%)	0042	1.3200	0.00000	1.3647	0.06623	0.03011	101	-0.68	032.42	0
032	Potassium (%)	0693	1.3239	0.00830	1.3647	0.06623	0.03011	101	-0.62	032.42	0
032	Potassium (%)	0563	1.3248	0.04630	1.3647	0.06623	0.03011	101	-0.60	032.31	0
032	Potassium (%)	0098	1.3250	0.07000	1.3647	0.06623	0.03011	101	-0.60	032.41	0
032	Potassium (%)	0154	1.3296	0.01790	1.3647	0.06623	0.03011	101	-0.53	032.52	0
032	Potassium (%)	0650	1.3300	0.14000	1.3647	0.06623	0.03011	101	-0.52	032.31	0
032	Potassium (%)	0148	1.3300	0.02000	1.3647	0.06623	0.03011	101	-0.52	032.41	0
032	Potassium (%)	2012	1.3300	0.00000	1.3647	0.06623	0.03011	101	-0.52	032.41	0
032	Potassium (%)	0142	1.3317	0.00520	1.3647	0.06623	0.03011	101	-0.50	032.31	0
032	Potassium (%)	2055	1.3320	0.00290	1.3647	0.06623	0.03011	101	-0.50	032.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
032	Potassium (%)	0918	1.3325	0.02900	1.3647	0.06623	0.03011	101	-0.49	032.43	0
032	Potassium (%)	0100	1.3350	0.01000	1.3647	0.06623	0.03011	101	-0.45	032.99	0
032	Potassium (%)	0014	1.3400	0.04000	1.3647	0.06623	0.03011	101	-0.37	032.31	0
032	Potassium (%)	0144	1.3400	0.04000	1.3647	0.06623	0.03011	101	-0.37	032.41	0
032	Potassium (%)	0009	1.3401	0.01260	1.3647	0.06623	0.03011	101	-0.37	032.42	0
032	Potassium (%)	2146	1.3420	0.01400	1.3647	0.06623	0.03011	101	-0.34	032.51	0
032	Potassium (%)	0045	1.3450	0.09000	1.3647	0.06623	0.03011	101	-0.30	032.42	0
032	Potassium (%)	0590	1.3450	0.01000	1.3647	0.06623	0.03011	101	-0.30	032.99	0
032	Potassium (%)	0948	1.3475	0.00100	1.3647	0.06623	0.03011	101	-0.26	032.31	0
032	Potassium (%)	0006	1.3495	0.00300	1.3647	0.06623	0.03011	101	-0.23	032.43	0
032	Potassium (%)	0504	1.3500	0.00000	1.3647	0.06623	0.03011	101	-0.22	032.02	0
032	Potassium (%)	0139	1.3515	0.01700	1.3647	0.06623	0.03011	101	-0.20	032.41	0
032	Potassium (%)	0186	1.3550	0.01000	1.3647	0.06623	0.03011	101	-0.15	032.52	0
032	Potassium (%)	0560	1.3578	0.00500	1.3647	0.06623	0.03011	101	-0.10	032.42	0
032	Potassium (%)	0175	1.3600	0.04000	1.3647	0.06623	0.03011	101	-0.07	032.31	0
032	Potassium (%)	0619	1.3600	0.04000	1.3647	0.06623	0.03011	101	-0.07	032.41	0
032	Potassium (%)	0205	1.3640	0.02000	1.3647	0.06623	0.03011	101	-0.01	032.42	0
032	Potassium (%)	0049	1.3650	0.03000	1.3647	0.06623	0.03011	101	0.00	032.41	0
032	Potassium (%)	0955	1.3650	0.01000	1.3647	0.06623	0.03011	101	0.00	032.44	0
032	Potassium (%)	0881	1.3670	0.01400	1.3647	0.06623	0.03011	101	0.03	032.43	0
032	Potassium (%)	0683	1.3700	0.10000	1.3647	0.06623	0.03011	101	0.08	032.31	0
032	Potassium (%)	0675	1.3700	0.02000	1.3647	0.06623	0.03011	101	0.08	032.43	0
032	Potassium (%)	0870	1.3720	0.00200	1.3647	0.06623	0.03011	101	0.11	032.42	0
032	Potassium (%)	0034	1.3746	0.06350	1.3647	0.06623	0.03011	101	0.15	032.43	0
032	Potassium (%)	0169	1.3750	0.01000	1.3647	0.06623	0.03011	101	0.15	032.32	0
032	Potassium (%)	0164	1.3750	0.05000	1.3647	0.06623	0.03011	101	0.15	032.41	0
032	Potassium (%)	0353	1.3750	0.05000	1.3647	0.06623	0.03011	101	0.15	032.43	0
032	Potassium (%)	0354	1.3800	0.02000	1.3647	0.06623	0.03011	101	0.23	032.31	0
032	Potassium (%)	2129	1.3800	0.02000	1.3647	0.06623	0.03011	101	0.23	032.42	0
032	Potassium (%)	0510	1.3800	0.00000	1.3647	0.06623	0.03011	101	0.23	032.43	0
032	Potassium (%)	0294	1.3850	0.01000	1.3647	0.06623	0.03011	101	0.31	032.42	0
032	Potassium (%)	0043	1.3850	0.01000	1.3647	0.06623	0.03011	101	0.31	032.43	0
032	Potassium (%)	0035	1.3850	0.05000	1.3647	0.06623	0.03011	101	0.31	032.42	0
032	Potassium (%)	0037	1.3850	0.03000	1.3647	0.06623	0.03011	101	0.31	032.42	0
032	Potassium (%)	0083	1.3850	0.05000	1.3647	0.06623	0.03011	101	0.31	032.43	0
032	Potassium (%)	0610	1.3900	0.02000	1.3647	0.06623	0.03011	101	0.38	032.43	0
032	Potassium (%)	0413	1.3900	0.04000	1.3647	0.06623	0.03011	101	0.38	032.41	0
032	Potassium (%)	0553	1.3950	0.01000	1.3647	0.06623	0.03011	101	0.46	032.53	0
032	Potassium (%)	0074	1.4000	0.00000	1.3647	0.06623	0.03011	101	0.53	032.41	0
032	Potassium (%)	0878	1.4000	0.00000	1.3647	0.06623	0.03011	101	0.53	032.41	0
032	Potassium (%)	0027	1.4015	0.04100	1.3647	0.06623	0.03011	101	0.55	032.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
032	Potassium (%)	0504	1.4040	0.04000	1.3647	0.06623	0.03011	101	0.59	032.42	0
032	Potassium (%)	0187	1.4050	0.01000	1.3647	0.06623	0.03011	101	0.61	032.42	0
032	Potassium (%)	0010	1.4050	0.03000	1.3647	0.06623	0.03011	101	0.61	032.43	0
032	Potassium (%)	0208	1.4100	0.04000	1.3647	0.06623	0.03011	101	0.68	032.31	0
032	Potassium (%)	0226	1.4100	0.02000	1.3647	0.06623	0.03011	101	0.68	032.41	0
032	Potassium (%)	0036	1.4137	0.00030	1.3647	0.06623	0.03011	101	0.74	032.43	0
032	Potassium (%)	0520	1.4150	0.01000	1.3647	0.06623	0.03011	101	0.76	032.41	0
032	Potassium (%)	0186	1.4200	0.02000	1.3647	0.06623	0.03011	101	0.83	032.42	0
032	Potassium (%)	2089	1.4200	0.00000	1.3647	0.06623	0.03011	101	0.83	032.43	0
032	Potassium (%)	2053	1.4250	0.01000	1.3647	0.06623	0.03011	101	0.91	032.31	0
032	Potassium (%)	0190	1.4350	0.01000	1.3647	0.06623	0.03011	101	1.06	032.42	0
032	Potassium (%)	0425	1.4350	0.01000	1.3647	0.06623	0.03011	101	1.06	032.43	0
032	Potassium (%)	0726	1.4365	0.00100	1.3647	0.06623	0.03011	101	1.08	032.42	0
032	Potassium (%)	0407	1.4425	0.04630	1.3647	0.06623	0.03011	101	1.17	032.41	0
032	Potassium (%)	0035	1.4550	0.01000	1.3647	0.06623	0.03011	101	1.36	032.32	0
032	Potassium (%)	0265	1.4550	0.05000	1.3647	0.06623	0.03011	101	1.36	032.42	0
032	Potassium (%)	2004	1.4550	0.01000	1.3647	0.06623	0.03011	101	1.36	032.44	0
032	Potassium (%)	0598	1.4565	0.02900	1.3647	0.06623	0.03011	101	1.39	032.41	0
032	Potassium (%)	0572	1.4800	0.06000	1.3647	0.06623	0.03011	101	1.74	032.53	0
032	Potassium (%)	0035	1.4800	0.02000	1.3647	0.06623	0.03011	101	1.74	032.99	0
032	Potassium (%)	0357	1.5106	0.01030	1.3647	0.06623	0.03011	101	2.20	032.42	0
032	Potassium (%)	0366	1.5150	0.03000	1.3647	0.06623	0.03011	101	2.27	032.42	0
032	Potassium (%)	0405	1.5550	0.01000	1.3647	0.06623	0.03011	101	2.87	032.41	0
032	Potassium (%)	0358	1.5700	0.08000	1.3647	0.06623	0.03011	101	3.10	032.41	0
032	Potassium (%)	0682	1.5900	0.00000	1.3647	0.06623	0.03011	101	3.40	032.43	0
032	Potassium (%)	0692	1.6550	0.07000	1.3647	0.06623	0.03011	101	4.38	032.42	0
032	Potassium (%)	0536	2.1500	0.16000	1.3647	0.06623	0.03011	101	11.86	032.31	0
032	Potassium (%)	0297	1.2900	0.22000	1.3647	0.06623	0.03011	101	-1.13	032.43	1
032	Potassium (%)	0208	1.4350	0.25600	1.3647	0.06623	0.03011	101	1.06	032.41	1
032	Potassium (%)	0171	2.3650	0.01000	1.3647	0.06623	0.03011	101	15.10	032.41	2
033	Salt as chloride (%)	0536	1.2700	0.04000	1.5893	0.06149	0.01917	66	-5.19	033.99	0
033	Salt as chloride (%)	0893	1.4100	0.00000	1.5893	0.06149	0.01917	66	-2.92	033.00	0
033	Salt as chloride (%)	0619	1.4150	0.03000	1.5893	0.06149	0.01917	66	-2.83	033.01	0
033	Salt as chloride (%)	0948	1.4200	0.12000	1.5893	0.06149	0.01917	66	-2.75	033.01	0
033	Salt as chloride (%)	0358	1.4400	0.02000	1.5893	0.06149	0.01917	66	-2.43	033.99	0
033	Salt as chloride (%)	0353	1.4550	0.03000	1.5893	0.06149	0.01917	66	-2.18	033.00	0
033	Salt as chloride (%)	0895	1.4950	0.01000	1.5893	0.06149	0.01917	66	-1.53	033.00	0
033	Salt as chloride (%)	2022	1.5045	0.00100	1.5893	0.06149	0.01917	66	-1.38	033.00	0
033	Salt as chloride (%)	2093	1.5050	0.01000	1.5893	0.06149	0.01917	66	-1.37	033.00	0
033	Salt as chloride (%)	0885	1.5097	0.05900	1.5893	0.06149	0.01917	66	-1.29	033.00	0
033	Salt as chloride (%)	0505	1.5150	0.01000	1.5893	0.06149	0.01917	66	-1.21	033.03	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
033	Salt as chloride (%)	0910	1.5200	0.00000	1.5893	0.06149	0.01917	66	-1.13	033.99	0
033	Salt as chloride (%)	0903	1.5250	0.01000	1.5893	0.06149	0.01917	66	-1.05	033.00	0
033	Salt as chloride (%)	0552	1.5250	0.05000	1.5893	0.06149	0.01917	66	-1.05	033.99	0
033	Salt as chloride (%)	0675	1.5300	0.02000	1.5893	0.06149	0.01917	66	-0.96	033.00	0
033	Salt as chloride (%)	0045	1.5400	0.02000	1.5893	0.06149	0.01917	66	-0.80	033.00	0
033	Salt as chloride (%)	0878	1.5400	0.02000	1.5893	0.06149	0.01917	66	-0.80	033.00	0
033	Salt as chloride (%)	0298	1.5500	0.02000	1.5893	0.06149	0.01917	66	-0.64	033.00	0
033	Salt as chloride (%)	0425	1.5550	0.03000	1.5893	0.06149	0.01917	66	-0.56	033.01	0
033	Salt as chloride (%)	2062	1.5645	0.01820	1.5893	0.06149	0.01917	66	-0.40	033.00	0
033	Salt as chloride (%)	0921	1.5650	0.03000	1.5893	0.06149	0.01917	66	-0.39	033.00	0
033	Salt as chloride (%)	0897	1.5750	0.01000	1.5893	0.06149	0.01917	66	-0.23	033.00	0
033	Salt as chloride (%)	0504	1.5800	0.02000	1.5893	0.06149	0.01917	66	-0.15	033.00	0
033	Salt as chloride (%)	0511	1.5800	0.00000	1.5893	0.06149	0.01917	66	-0.15	033.00	0
033	Salt as chloride (%)	0689	1.5800	0.00000	1.5893	0.06149	0.01917	66	-0.15	033.05	0
033	Salt as chloride (%)	2009	1.5826	0.00070	1.5893	0.06149	0.01917	66	-0.11	033.00	0
033	Salt as chloride (%)	0354	1.5850	0.01000	1.5893	0.06149	0.01917	66	-0.07	033.01	0
033	Salt as chloride (%)	0004	1.5900	0.00000	1.5893	0.06149	0.01917	66	0.01	033.01	0
033	Salt as chloride (%)	0144	1.5950	0.13000	1.5893	0.06149	0.01917	66	0.09	033.03	0
033	Salt as chloride (%)	0194	1.5950	0.01000	1.5893	0.06149	0.01917	66	0.09	033.01	0
033	Salt as chloride (%)	0650	1.5950	0.03000	1.5893	0.06149	0.01917	66	0.09	033.01	0
033	Salt as chloride (%)	0510	1.6000	0.00000	1.5893	0.06149	0.01917	66	0.17	033.01	0
033	Salt as chloride (%)	0874	1.6000	0.00000	1.5893	0.06149	0.01917	66	0.17	033.01	0
033	Salt as chloride (%)	0190	1.6000	0.08000	1.5893	0.06149	0.01917	66	0.17	033.03	0
033	Salt as chloride (%)	2145	1.6050	0.03000	1.5893	0.06149	0.01917	66	0.26	033.00	0
033	Salt as chloride (%)	0171	1.6050	0.01000	1.5893	0.06149	0.01917	66	0.26	033.05	0
033	Salt as chloride (%)	2068	1.6100	0.00000	1.5893	0.06149	0.01917	66	0.34	033.00	0
033	Salt as chloride (%)	0644	1.6105	0.00620	1.5893	0.06149	0.01917	66	0.35	033.01	0
033	Salt as chloride (%)	2146	1.6150	0.03000	1.5893	0.06149	0.01917	66	0.42	033.01	0
033	Salt as chloride (%)	0683	1.6150	0.01000	1.5893	0.06149	0.01917	66	0.42	033.01	0
033	Salt as chloride (%)	0309	1.6188	0.00210	1.5893	0.06149	0.01917	66	0.48	033.00	0
033	Salt as chloride (%)	0539	1.6200	0.04000	1.5893	0.06149	0.01917	66	0.50	033.00	0
033	Salt as chloride (%)	0098	1.6200	0.02000	1.5893	0.06149	0.01917	66	0.50	033.01	0
033	Salt as chloride (%)	0100	1.6200	0.00000	1.5893	0.06149	0.01917	66	0.50	033.01	0
033	Salt as chloride (%)	2109	1.6200	0.00400	1.5893	0.06149	0.01917	66	0.50	033.01	0
033	Salt as chloride (%)	0881	1.6205	0.00900	1.5893	0.06149	0.01917	66	0.51	033.00	0
033	Salt as chloride (%)	0175	1.6250	0.01000	1.5893	0.06149	0.01917	66	0.58	033.01	0
033	Salt as chloride (%)	0366	1.6300	0.02000	1.5893	0.06149	0.01917	66	0.66	033.00	0
033	Salt as chloride (%)	0001	1.6300	0.02000	1.5893	0.06149	0.01917	66	0.66	033.01	0
033	Salt as chloride (%)	0590	1.6300	0.02000	1.5893	0.06149	0.01917	66	0.66	033.01	0
033	Salt as chloride (%)	0861	1.6300	0.00000	1.5893	0.06149	0.01917	66	0.66	033.01	0
033	Salt as chloride (%)	2073	1.6300	0.02000	1.5893	0.06149	0.01917	66	0.66	033.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
033	Salt as chloride (%)	0148	1.6355	0.00300	1.5893	0.06149	0.01917	66	0.75	033.01	0
033	Salt as chloride (%)	0726	1.6400	0.00000	1.5893	0.06149	0.01917	66	0.82	033.03	0
033	Salt as chloride (%)	0407	1.6430	0.00200	1.5893	0.06149	0.01917	66	0.87	033.01	0
033	Salt as chloride (%)	0019	1.6450	0.01000	1.5893	0.06149	0.01917	66	0.91	033.01	0
033	Salt as chloride (%)	0629	1.6450	0.01000	1.5893	0.06149	0.01917	66	0.91	033.01	0
033	Salt as chloride (%)	0208	1.6450	0.03000	1.5893	0.06149	0.01917	66	0.91	033.99	0
033	Salt as chloride (%)	0169	1.6500	0.00000	1.5893	0.06149	0.01917	66	0.99	033.00	0
033	Salt as chloride (%)	0164	1.6500	0.00000	1.5893	0.06149	0.01917	66	0.99	033.01	0
033	Salt as chloride (%)	0226	1.6500	0.04000	1.5893	0.06149	0.01917	66	0.99	033.01	0
033	Salt as chloride (%)	0229	1.6500	0.00000	1.5893	0.06149	0.01917	66	0.99	033.01	0
033	Salt as chloride (%)	0278	1.6550	0.01000	1.5893	0.06149	0.01917	66	1.07	033.01	0
033	Salt as chloride (%)	0682	1.6800	0.00000	1.5893	0.06149	0.01917	66	1.48	033.99	0
033	Salt as chloride (%)	0205	1.6900	0.02000	1.5893	0.06149	0.01917	66	1.64	033.01	0
033	Salt as chloride (%)	0265	1.7450	0.05000	1.5893	0.06149	0.01917	66	2.53	033.03	0
033	Salt as chloride (%)	0693	1.7429	0.17700	1.5893	0.06149	0.01917	66	2.50	033.00	1
033	Salt as chloride (%)	0083	1.7500	0.32000	1.5893	0.06149	0.01917	66	2.61	033.01	1
034	Selenium (mg / kg (ppm))	0045	0.32050	0.05300	1.9126	0.33560	0.08000	21	-4.74	034.04	0
034	Selenium (mg / kg (ppm))	0939	1.1750	0.11000	1.9126	0.33560	0.08000	21	-2.20	034.04	0
034	Selenium (mg / kg (ppm))	0675	1.2950	0.01000	1.9126	0.33560	0.08000	21	-1.84	034.43	0
034	Selenium (mg / kg (ppm))	0171	1.6650	0.05000	1.9126	0.33560	0.08000	21	-0.74	034.04	0
034	Selenium (mg / kg (ppm))	2055	1.6650	0.33000	1.9126	0.33560	0.08000	21	-0.74	034.53	0
034	Selenium (mg / kg (ppm))	0148	1.7050	0.01000	1.9126	0.33560	0.08000	21	-0.62	034.41	0
034	Selenium (mg / kg (ppm))	0563	1.7434	0.02430	1.9126	0.33560	0.08000	21	-0.50	034.04	0
034	Selenium (mg / kg (ppm))	0619	1.7500	0.28000	1.9126	0.33560	0.08000	21	-0.48	034.41	0
034	Selenium (mg / kg (ppm))	0692	1.8000	0.20000	1.9126	0.33560	0.08000	21	-0.34	034.42	0
034	Selenium (mg / kg (ppm))	0186	1.9100	0.06000	1.9126	0.33560	0.08000	21	-0.01	034.52	0
034	Selenium (mg / kg (ppm))	0169	2.0500	0.04000	1.9126	0.33560	0.08000	21	0.41	034.04	0
034	Selenium (mg / kg (ppm))	0610	2.0500	0.02000	1.9126	0.33560	0.08000	21	0.41	034.04	0
034	Selenium (mg / kg (ppm))	0047	2.0650	0.03000	1.9126	0.33560	0.08000	21	0.45	034.52	0
034	Selenium (mg / kg (ppm))	0164	2.0850	0.01000	1.9126	0.33560	0.08000	21	0.51	034.04	0
034	Selenium (mg / kg (ppm))	0190	2.0900	0.00000	1.9126	0.33560	0.08000	21	0.53	034.99	0
034	Selenium (mg / kg (ppm))	0098	2.0950	0.02800	1.9126	0.33560	0.08000	21	0.54	034.53	0
034	Selenium (mg / kg (ppm))	0553	2.1250	0.07000	1.9126	0.33560	0.08000	21	0.63	034.53	0
034	Selenium (mg / kg (ppm))	0918	2.2390	0.26380	1.9126	0.33560	0.08000	21	0.97	034.53	0
034	Selenium (mg / kg (ppm))	0910	2.2900	0.06000	1.9126	0.33560	0.08000	21	1.12	034.52	0
034	Selenium (mg / kg (ppm))	0208	2.2950	0.01000	1.9126	0.33560	0.08000	21	1.14	034.52	0
034	Selenium (mg / kg (ppm))	0964	2.3155	0.02100	1.9126	0.33560	0.08000	21	1.20	034.43	0
035	Sodium (%)	0511	0.22500	0.01000	0.49655	0.03064	0.01501	99	-8.86	035.41	0
035	Sodium (%)	0675	0.38500	0.01000	0.49655	0.03064	0.01501	99	-3.64	035.43	0
035	Sodium (%)	0175	0.42500	0.01000	0.49655	0.03064	0.01501	99	-2.34	035.31	0
035	Sodium (%)	0233	0.43500	0.01000	0.49655	0.03064	0.01501	99	-2.01	035.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
035	Sodium (%)	0689	0.45000	0.00000	0.49655	0.03064	0.01501	99	-1.52	035.31	0
035	Sodium (%)	0510	0.45450	0.00100	0.49655	0.03064	0.01501	99	-1.37	035.43	0
035	Sodium (%)	0910	0.45500	0.01000	0.49655	0.03064	0.01501	99	-1.36	035.41	0
035	Sodium (%)	0142	0.45625	0.00650	0.49655	0.03064	0.01501	99	-1.32	035.31	0
035	Sodium (%)	0298	0.46000	0.02000	0.49655	0.03064	0.01501	99	-1.19	035.41	0
035	Sodium (%)	0366	0.46000	0.02000	0.49655	0.03064	0.01501	99	-1.19	035.42	0
035	Sodium (%)	0033	0.46000	0.04000	0.49655	0.03064	0.01501	99	-1.19	035.43	0
035	Sodium (%)	0035	0.46000	0.02000	0.49655	0.03064	0.01501	99	-1.19	035.43	0
035	Sodium (%)	0650	0.46500	0.01000	0.49655	0.03064	0.01501	99	-1.03	035.31	0
035	Sodium (%)	0098	0.46500	0.03000	0.49655	0.03064	0.01501	99	-1.03	035.43	0
035	Sodium (%)	0202	0.46500	0.01000	0.49655	0.03064	0.01501	99	-1.03	035.43	0
035	Sodium (%)	0563	0.46510	0.01940	0.49655	0.03064	0.01501	99	-1.03	035.31	0
035	Sodium (%)	0683	0.47000	0.00000	0.49655	0.03064	0.01501	99	-0.87	035.31	0
035	Sodium (%)	0208	0.47050	0.03500	0.49655	0.03064	0.01501	99	-0.85	035.41	0
035	Sodium (%)	0874	0.47200	0.05400	0.49655	0.03064	0.01501	99	-0.80	035.31	0
035	Sodium (%)	0861	0.47300	0.02000	0.49655	0.03064	0.01501	99	-0.77	035.43	0
035	Sodium (%)	0629	0.47350	0.00700	0.49655	0.03064	0.01501	99	-0.75	035.31	0
035	Sodium (%)	0205	0.47400	0.01200	0.49655	0.03064	0.01501	99	-0.74	035.42	0
035	Sodium (%)	0144	0.47500	0.01000	0.49655	0.03064	0.01501	99	-0.70	035.41	0
035	Sodium (%)	2022	0.47600	0.00600	0.49655	0.03064	0.01501	99	-0.67	035.31	0
035	Sodium (%)	2006	0.48000	0.00200	0.49655	0.03064	0.01501	99	-0.54	035.01	0
035	Sodium (%)	0939	0.48000	0.00000	0.49655	0.03064	0.01501	99	-0.54	035.31	0
035	Sodium (%)	0148	0.48000	0.02000	0.49655	0.03064	0.01501	99	-0.54	035.41	0
035	Sodium (%)	2012	0.48000	0.00000	0.49655	0.03064	0.01501	99	-0.54	035.41	0
035	Sodium (%)	0297	0.48000	0.06000	0.49655	0.03064	0.01501	99	-0.54	035.43	0
035	Sodium (%)	0629	0.48000	0.00000	0.49655	0.03064	0.01501	99	-0.54	035.43	0
035	Sodium (%)	0505	0.48150	0.00700	0.49655	0.03064	0.01501	99	-0.49	035.31	0
035	Sodium (%)	0345	0.48150	0.00300	0.49655	0.03064	0.01501	99	-0.49	035.43	0
035	Sodium (%)	0019	0.48500	0.01000	0.49655	0.03064	0.01501	99	-0.38	035.41	0
035	Sodium (%)	0529	0.48550	0.00500	0.49655	0.03064	0.01501	99	-0.36	035.31	0
035	Sodium (%)	0881	0.48600	0.00200	0.49655	0.03064	0.01501	99	-0.34	035.43	0
035	Sodium (%)	0164	0.48700	0.00200	0.49655	0.03064	0.01501	99	-0.31	035.41	0
035	Sodium (%)	0520	0.48725	0.00870	0.49655	0.03064	0.01501	99	-0.30	035.41	0
035	Sodium (%)	0870	0.48735	0.00250	0.49655	0.03064	0.01501	99	-0.30	035.42	0
035	Sodium (%)	0045	0.48750	0.04300	0.49655	0.03064	0.01501	99	-0.30	035.42	0
035	Sodium (%)	0560	0.48815	0.00450	0.49655	0.03064	0.01501	99	-0.27	035.42	0
035	Sodium (%)	0263	0.48830	0.00140	0.49655	0.03064	0.01501	99	-0.27	035.32	0
035	Sodium (%)	0918	0.48900	0.00600	0.49655	0.03064	0.01501	99	-0.25	035.43	0
035	Sodium (%)	0035	0.49000	0.04000	0.49655	0.03064	0.01501	99	-0.21	035.42	0
035	Sodium (%)	0186	0.49000	0.00600	0.49655	0.03064	0.01501	99	-0.21	035.42	0
035	Sodium (%)	0682	0.49000	0.00000	0.49655	0.03064	0.01501	99	-0.21	035.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
035	Sodium (%)	0590	0.49000	0.00000	0.49655	0.03064	0.01501	99	-0.21	035.99	0
035	Sodium (%)	0598	0.49220	0.00540	0.49655	0.03064	0.01501	99	-0.14	035.41	0
035	Sodium (%)	2146	0.49400	0.00400	0.49655	0.03064	0.01501	99	-0.08	035.51	0
035	Sodium (%)	0693	0.49460	0.00100	0.49655	0.03064	0.01501	99	-0.06	035.42	0
035	Sodium (%)	0413	0.49500	0.01000	0.49655	0.03064	0.01501	99	-0.05	035.41	0
035	Sodium (%)	0010	0.49500	0.01000	0.49655	0.03064	0.01501	99	-0.05	035.43	0
035	Sodium (%)	0955	0.49550	0.01300	0.49655	0.03064	0.01501	99	-0.03	035.41	0
035	Sodium (%)	0964	0.49800	0.04600	0.49655	0.03064	0.01501	99	0.05	035.43	0
035	Sodium (%)	0154	0.49835	0.01030	0.49655	0.03064	0.01501	99	0.06	035.52	0
035	Sodium (%)	0504	0.50000	0.00000	0.49655	0.03064	0.01501	99	0.11	035.05	0
035	Sodium (%)	0098	0.50000	0.02000	0.49655	0.03064	0.01501	99	0.11	035.41	0
035	Sodium (%)	0878	0.50000	0.02000	0.49655	0.03064	0.01501	99	0.11	035.41	0
035	Sodium (%)	0100	0.50000	0.00000	0.49655	0.03064	0.01501	99	0.11	035.99	0
035	Sodium (%)	2105	0.50000	0.00000	0.49655	0.03064	0.01501	99	0.11	035.99	0
035	Sodium (%)	0619	0.50050	0.01100	0.49655	0.03064	0.01501	99	0.13	035.41	0
035	Sodium (%)	2004	0.50100	0.01600	0.49655	0.03064	0.01501	99	0.15	035.41	0
035	Sodium (%)	0504	0.50150	0.00700	0.49655	0.03064	0.01501	99	0.16	035.42	0
035	Sodium (%)	0034	0.50215	0.01950	0.49655	0.03064	0.01501	99	0.18	035.43	0
035	Sodium (%)	0186	0.50250	0.01100	0.49655	0.03064	0.01501	99	0.19	035.52	0
035	Sodium (%)	0208	0.50350	0.04300	0.49655	0.03064	0.01501	99	0.23	035.31	0
035	Sodium (%)	0572	0.50350	0.03700	0.49655	0.03064	0.01501	99	0.23	035.53	0
035	Sodium (%)	0049	0.50500	0.01000	0.49655	0.03064	0.01501	99	0.28	035.41	0
035	Sodium (%)	0083	0.50500	0.01000	0.49655	0.03064	0.01501	99	0.28	035.43	0
035	Sodium (%)	0353	0.50500	0.07000	0.49655	0.03064	0.01501	99	0.28	035.43	0
035	Sodium (%)	0425	0.50500	0.01000	0.49655	0.03064	0.01501	99	0.28	035.43	0
035	Sodium (%)	0003	0.50950	0.01100	0.49655	0.03064	0.01501	99	0.42	035.05	0
035	Sodium (%)	2055	0.50970	0.00140	0.49655	0.03064	0.01501	99	0.43	035.53	0
035	Sodium (%)	0035	0.51000	0.00000	0.49655	0.03064	0.01501	99	0.44	035.32	0
035	Sodium (%)	0610	0.51000	0.02000	0.49655	0.03064	0.01501	99	0.44	035.43	0
035	Sodium (%)	0139	0.51500	0.00400	0.49655	0.03064	0.01501	99	0.60	035.41	0
035	Sodium (%)	0190	0.51500	0.01000	0.49655	0.03064	0.01501	99	0.60	035.42	0
035	Sodium (%)	0265	0.51500	0.01000	0.49655	0.03064	0.01501	99	0.60	035.42	0
035	Sodium (%)	0948	0.51900	0.01000	0.49655	0.03064	0.01501	99	0.73	035.31	0
035	Sodium (%)	0278	0.52000	0.04000	0.49655	0.03064	0.01501	99	0.77	035.42	0
035	Sodium (%)	2089	0.52000	0.00000	0.49655	0.03064	0.01501	99	0.77	035.43	0
035	Sodium (%)	0187	0.52320	0.00120	0.49655	0.03064	0.01501	99	0.87	035.42	0
035	Sodium (%)	0035	0.52500	0.07000	0.49655	0.03064	0.01501	99	0.93	035.99	0
035	Sodium (%)	0294	0.53000	0.00000	0.49655	0.03064	0.01501	99	1.09	035.42	0
035	Sodium (%)	0300	0.53600	0.03600	0.49655	0.03064	0.01501	99	1.29	035.53	0
035	Sodium (%)	0726	0.53670	0.00020	0.49655	0.03064	0.01501	99	1.31	035.42	0
035	Sodium (%)	0108	0.54000	0.04000	0.49655	0.03064	0.01501	99	1.42	035.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
035	Sodium (%)	0407	0.54040	0.00600	0.49655	0.03064	0.01501	99	1.43	035.41	0
035	Sodium (%)	0014	0.54450	0.08500	0.49655	0.03064	0.01501	99	1.56	035.31	0
035	Sodium (%)	0169	0.54500	0.01000	0.49655	0.03064	0.01501	99	1.58	035.32	0
035	Sodium (%)	0405	0.55500	0.01000	0.49655	0.03064	0.01501	99	1.91	035.41	0
035	Sodium (%)	0226	0.55650	0.00100	0.49655	0.03064	0.01501	99	1.96	035.41	0
035	Sodium (%)	0354	0.56500	0.01000	0.49655	0.03064	0.01501	99	2.23	035.31	0
035	Sodium (%)	0358	0.56500	0.01000	0.49655	0.03064	0.01501	99	2.23	035.41	0
035	Sodium (%)	0553	0.57300	0.03200	0.49655	0.03064	0.01501	99	2.49	035.53	0
035	Sodium (%)	0692	0.57500	0.05000	0.49655	0.03064	0.01501	99	2.56	035.42	0
035	Sodium (%)	0152	0.58500	0.01000	0.49655	0.03064	0.01501	99	2.89	035.05	0
035	Sodium (%)	0921	0.61800	0.00000	0.49655	0.03064	0.01501	99	3.96	035.31	0
035	Sodium (%)	0536	0.73500	0.03000	0.49655	0.03064	0.01501	99	7.78	035.31	0
035	Sodium (%)	0171	0.78000	0.00000	0.49655	0.03064	0.01501	99	9.25	035.41	0
035	Sodium (%)	0042	0.56500	0.15600	0.49655	0.03064	0.01501	99	2.23	035.42	1
036	Sulfur (%)	0098	0.23500	0.01000	0.29834	0.02159	0.01062	33	-2.93	036.04	0
036	Sulfur (%)	2053	0.26500	0.01000	0.29834	0.02159	0.01062	33	-1.54	036.04	0
036	Sulfur (%)	0035	0.27500	0.01000	0.29834	0.02159	0.01062	33	-1.08	036.43	0
036	Sulfur (%)	0353	0.27500	0.01000	0.29834	0.02159	0.01062	33	-1.08	036.43	0
036	Sulfur (%)	2089	0.27500	0.03000	0.29834	0.02159	0.01062	33	-1.08	036.43	0
036	Sulfur (%)	0186	0.28350	0.00100	0.29834	0.02159	0.01062	33	-0.69	036.52	0
036	Sulfur (%)	0366	0.28500	0.01000	0.29834	0.02159	0.01062	33	-0.62	036.42	0
036	Sulfur (%)	0202	0.28500	0.01000	0.29834	0.02159	0.01062	33	-0.62	036.43	0
036	Sulfur (%)	0297	0.28500	0.05000	0.29834	0.02159	0.01062	33	-0.62	036.43	0
036	Sulfur (%)	0870	0.28655	0.01150	0.29834	0.02159	0.01062	33	-0.55	036.42	0
036	Sulfur (%)	0560	0.28660	0.00020	0.29834	0.02159	0.01062	33	-0.54	036.42	0
036	Sulfur (%)	0407	0.28775	0.01050	0.29834	0.02159	0.01062	33	-0.49	036.42	0
036	Sulfur (%)	0629	0.28850	0.00300	0.29834	0.02159	0.01062	33	-0.46	036.43	0
036	Sulfur (%)	0045	0.28900	0.00000	0.29834	0.02159	0.01062	33	-0.43	036.42	0
036	Sulfur (%)	0265	0.29000	0.02000	0.29834	0.02159	0.01062	33	-0.39	036.42	0
036	Sulfur (%)	0693	0.29010	0.00540	0.29834	0.02159	0.01062	33	-0.38	036.42	0
036	Sulfur (%)	0035	0.29500	0.01000	0.29834	0.02159	0.01062	33	-0.15	036.42	0
036	Sulfur (%)	0164	0.29500	0.01000	0.29834	0.02159	0.01062	33	-0.15	036.42	0
036	Sulfur (%)	0186	0.29600	0.01200	0.29834	0.02159	0.01062	33	-0.11	036.42	0
036	Sulfur (%)	0345	0.30300	0.00200	0.29834	0.02159	0.01062	33	0.22	036.43	0
036	Sulfur (%)	0910	0.30550	0.00000	0.29834	0.02159	0.01062	33	0.33	036.52	0
036	Sulfur (%)	0294	0.31000	0.00000	0.29834	0.02159	0.01062	33	0.54	036.42	0
036	Sulfur (%)	0918	0.31150	0.00700	0.29834	0.02159	0.01062	33	0.61	036.43	0
036	Sulfur (%)	0187	0.31200	0.00120	0.29834	0.02159	0.01062	33	0.63	036.42	0
036	Sulfur (%)	0357	0.31265	0.00470	0.29834	0.02159	0.01062	33	0.66	036.42	0
036	Sulfur (%)	0098	0.31430	0.03400	0.29834	0.02159	0.01062	33	0.74	036.43	0
036	Sulfur (%)	0278	0.31500	0.03000	0.29834	0.02159	0.01062	33	0.77	036.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
036	Sulfur (%)	0553	0.31850	0.01100	0.29834	0.02159	0.01062	33	0.93	036.53	0
036	Sulfur (%)	0510	0.32000	0.00000	0.29834	0.02159	0.01062	33	1.00	036.43	0
036	Sulfur (%)	0708	0.33350	0.01100	0.29834	0.02159	0.01062	33	1.63	036.42	0
036	Sulfur (%)	0226	0.35000	0.02000	0.29834	0.02159	0.01062	33	2.39	036.04	0
036	Sulfur (%)	0692	0.35000	0.00000	0.29834	0.02159	0.01062	33	2.39	036.42	0
036	Sulfur (%)	0171	0.61300	0.00600	0.29834	0.02159	0.01062	33	14.57	036.42	0
036	Sulfur (%)	0037	0.32500	0.07000	0.29834	0.02159	0.01062	33	1.23	036.42	1
036	Sulfur (%)	0169	0.00000	0.00000	0.29834	0.02159	0.01062	33		036.43	4
037	Zinc (mg / kg (ppm))	0536	99.950	0.40000	3,484.4	240.05	75.037	98	-14.10	037.31	0
037	Zinc (mg / kg (ppm))	0026	245.70	4.2000	3,484.4	240.05	75.037	98	-13.49	037.42	0
037	Zinc (mg / kg (ppm))	0171	1,145.0	10.000	3,484.4	240.05	75.037	98	-9.75	037.41	0
037	Zinc (mg / kg (ppm))	0921	1,625.2	0.00000	3,484.4	240.05	75.037	98	-7.75	037.31	0
037	Zinc (mg / kg (ppm))	0405	2,292.0	4.0000	3,484.4	240.05	75.037	98	-4.97	037.41	0
037	Zinc (mg / kg (ppm))	0047	2,964.8	0.00000	3,484.4	240.05	75.037	98	-2.16	037.52	0
037	Zinc (mg / kg (ppm))	0045	3,000.0	100.00	3,484.4	240.05	75.037	98	-2.02	037.42	0
037	Zinc (mg / kg (ppm))	0861	3,005.0	10.000	3,484.4	240.05	75.037	98	-2.00	037.43	0
037	Zinc (mg / kg (ppm))	0083	3,020.0	150.00	3,484.4	240.05	75.037	98	-1.93	037.43	0
037	Zinc (mg / kg (ppm))	0874	3,120.0	20.000	3,484.4	240.05	75.037	98	-1.52	037.31	0
037	Zinc (mg / kg (ppm))	0345	3,165.0	70.000	3,484.4	240.05	75.037	98	-1.33	037.43	0
037	Zinc (mg / kg (ppm))	0028	3,194.0	48.000	3,484.4	240.05	75.037	98	-1.21	037.43	0
037	Zinc (mg / kg (ppm))	0034	3,200.3	162.10	3,484.4	240.05	75.037	98	-1.18	037.43	0
037	Zinc (mg / kg (ppm))	0001	3,221.5	119.00	3,484.4	240.05	75.037	98	-1.10	037.31	0
037	Zinc (mg / kg (ppm))	0870	3,230.0	78.000	3,484.4	240.05	75.037	98	-1.06	037.42	0
037	Zinc (mg / kg (ppm))	0033	3,245.0	230.00	3,484.4	240.05	75.037	98	-1.00	037.43	0
037	Zinc (mg / kg (ppm))	0278	3,258.6	145.20	3,484.4	240.05	75.037	98	-0.94	037.42	0
037	Zinc (mg / kg (ppm))	0510	3,268.5	17.000	3,484.4	240.05	75.037	98	-0.90	037.43	0
037	Zinc (mg / kg (ppm))	0358	3,273.6	140.74	3,484.4	240.05	75.037	98	-0.88	037.41	0
037	Zinc (mg / kg (ppm))	2055	3,279.0	346.00	3,484.4	240.05	75.037	98	-0.86	037.53	0
037	Zinc (mg / kg (ppm))	0955	3,280.5	69.000	3,484.4	240.05	75.037	98	-0.85	037.44	0
037	Zinc (mg / kg (ppm))	0511	3,312.5	51.000	3,484.4	240.05	75.037	98	-0.72	037.41	0
037	Zinc (mg / kg (ppm))	0563	3,318.9	7.5000	3,484.4	240.05	75.037	98	-0.69	037.31	0
037	Zinc (mg / kg (ppm))	0629	3,320.0	0.00000	3,484.4	240.05	75.037	98	-0.69	037.43	0
037	Zinc (mg / kg (ppm))	0098	3,329.0	214.00	3,484.4	240.05	75.037	98	-0.65	037.41	0
037	Zinc (mg / kg (ppm))	0610	3,329.0	26.000	3,484.4	240.05	75.037	98	-0.65	037.43	0
037	Zinc (mg / kg (ppm))	0187	3,342.2	0.87000	3,484.4	240.05	75.037	98	-0.59	037.42	0
037	Zinc (mg / kg (ppm))	0918	3,361.9	48.140	3,484.4	240.05	75.037	98	-0.51	037.43	0
037	Zinc (mg / kg (ppm))	0294	3,370.0	11.280	3,484.4	240.05	75.037	98	-0.48	037.42	0
037	Zinc (mg / kg (ppm))	0560	3,371.0	83.400	3,484.4	240.05	75.037	98	-0.47	037.42	0
037	Zinc (mg / kg (ppm))	2068	3,377.1	5.7130	3,484.4	240.05	75.037	98	-0.45	037.31	0
037	Zinc (mg / kg (ppm))	2022	3,383.5	37.000	3,484.4	240.05	75.037	98	-0.42	037.31	0
037	Zinc (mg / kg (ppm))	0413	3,385.0	230.00	3,484.4	240.05	75.037	98	-0.41	037.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
037	Zinc (mg / kg (ppm))	0265	3,407.5	193.00	3,484.4	240.05	75.037	98	-0.32	037.42	0
037	Zinc (mg / kg (ppm))	0529	3,418.8	175.00	3,484.4	240.05	75.037	98	-0.27	037.31	0
037	Zinc (mg / kg (ppm))	0629	3,420.0	40.000	3,484.4	240.05	75.037	98	-0.27	037.31	0
037	Zinc (mg / kg (ppm))	2089	3,424.5	91.630	3,484.4	240.05	75.037	98	-0.25	037.43	0
037	Zinc (mg / kg (ppm))	2004	3,425.0	30.000	3,484.4	240.05	75.037	98	-0.25	037.44	0
037	Zinc (mg / kg (ppm))	0572	3,425.0	130.00	3,484.4	240.05	75.037	98	-0.25	037.53	0
037	Zinc (mg / kg (ppm))	0689	3,437.1	31.600	3,484.4	240.05	75.037	98	-0.20	037.31	0
037	Zinc (mg / kg (ppm))	0590	3,453.0	104.00	3,484.4	240.05	75.037	98	-0.13	037.99	0
037	Zinc (mg / kg (ppm))	0043	3,464.0	110.00	3,484.4	240.05	75.037	98	-0.09	037.43	0
037	Zinc (mg / kg (ppm))	0010	3,465.0	30.000	3,484.4	240.05	75.037	98	-0.08	037.33	0
037	Zinc (mg / kg (ppm))	0186	3,468.0	54.000	3,484.4	240.05	75.037	98	-0.07	037.52	0
037	Zinc (mg / kg (ppm))	0693	3,476.8	48.712	3,484.4	240.05	75.037	98	-0.03	037.42	0
037	Zinc (mg / kg (ppm))	0964	3,480.0	44.000	3,484.4	240.05	75.037	98	-0.02	037.43	0
037	Zinc (mg / kg (ppm))	0598	3,480.5	65.000	3,484.4	240.05	75.037	98	-0.02	037.41	0
037	Zinc (mg / kg (ppm))	0202	3,482.7	27.520	3,484.4	240.05	75.037	98	-0.01	037.43	0
037	Zinc (mg / kg (ppm))	0504	3,486.5	37.000	3,484.4	240.05	75.037	98	0.01	037.32	0
037	Zinc (mg / kg (ppm))	0049	3,486.9	46.100	3,484.4	240.05	75.037	98	0.01	037.41	0
037	Zinc (mg / kg (ppm))	0164	3,490.0	40.000	3,484.4	240.05	75.037	98	0.02	037.41	0
037	Zinc (mg / kg (ppm))	0505	3,495.0	20.000	3,484.4	240.05	75.037	98	0.04	037.31	0
037	Zinc (mg / kg (ppm))	0148	3,515.0	14.000	3,484.4	240.05	75.037	98	0.13	037.41	0
037	Zinc (mg / kg (ppm))	0407	3,523.0	27.030	3,484.4	240.05	75.037	98	0.16	037.41	0
037	Zinc (mg / kg (ppm))	0098	3,523.5	89.000	3,484.4	240.05	75.037	98	0.16	037.43	0
037	Zinc (mg / kg (ppm))	0013	3,535.0	50.000	3,484.4	240.05	75.037	98	0.21	037.43	0
037	Zinc (mg / kg (ppm))	0682	3,536.0	0.00000	3,484.4	240.05	75.037	98	0.21	037.43	0
037	Zinc (mg / kg (ppm))	0014	3,537.0	88.000	3,484.4	240.05	75.037	98	0.22	037.31	0
037	Zinc (mg / kg (ppm))	0504	3,538.5	15.000	3,484.4	240.05	75.037	98	0.23	037.33	0
037	Zinc (mg / kg (ppm))	0357	3,548.3	65.610	3,484.4	240.05	75.037	98	0.27	037.42	0
037	Zinc (mg / kg (ppm))	0175	3,550.0	100.00	3,484.4	240.05	75.037	98	0.27	037.31	0
037	Zinc (mg / kg (ppm))	0186	3,553.0	126.00	3,484.4	240.05	75.037	98	0.29	037.42	0
037	Zinc (mg / kg (ppm))	0019	3,553.5	73.000	3,484.4	240.05	75.037	98	0.29	037.41	0
037	Zinc (mg / kg (ppm))	0939	3,563.1	16.100	3,484.4	240.05	75.037	98	0.33	037.31	0
037	Zinc (mg / kg (ppm))	0425	3,574.0	38.000	3,484.4	240.05	75.037	98	0.37	037.43	0
037	Zinc (mg / kg (ppm))	0035	3,581.5	69.000	3,484.4	240.05	75.037	98	0.40	037.32	0
037	Zinc (mg / kg (ppm))	0037	3,584.0	80.000	3,484.4	240.05	75.037	98	0.41	037.42	0
037	Zinc (mg / kg (ppm))	0878	3,586.0	30.000	3,484.4	240.05	75.037	98	0.42	037.41	0
037	Zinc (mg / kg (ppm))	0948	3,588.6	4.5130	3,484.4	240.05	75.037	98	0.43	037.33	0
037	Zinc (mg / kg (ppm))	0009	3,595.4	83.162	3,484.4	240.05	75.037	98	0.46	037.42	0
037	Zinc (mg / kg (ppm))	0208	3,600.0	100.00	3,484.4	240.05	75.037	98	0.48	037.31	0
037	Zinc (mg / kg (ppm))	0653	3,605.5	13.000	3,484.4	240.05	75.037	98	0.50	037.31	0
037	Zinc (mg / kg (ppm))	2149	3,646.0	17.560	3,484.4	240.05	75.037	98	0.67	037.31	0
037	Zinc (mg / kg (ppm))	2145	3,652.0	24.000	3,484.4	240.05	75.037	98	0.70	037.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
037	Zinc (mg / kg (ppm))	0190	3,656.0	12.000	3,484.4	240.05	75.037	98	0.71	037.42	0
037	Zinc (mg / kg (ppm))	0208	3,659.5	189.00	3,484.4	240.05	75.037	98	0.73	037.41	0
037	Zinc (mg / kg (ppm))	0074	3,662.0	28.000	3,484.4	240.05	75.037	98	0.74	037.41	0
037	Zinc (mg / kg (ppm))	2012	3,664.3	1.5700	3,484.4	240.05	75.037	98	0.75	037.41	0
037	Zinc (mg / kg (ppm))	2129	3,682.0	138.00	3,484.4	240.05	75.037	98	0.82	037.99	0
037	Zinc (mg / kg (ppm))	0520	3,686.5	93.000	3,484.4	240.05	75.037	98	0.84	037.41	0
037	Zinc (mg / kg (ppm))	0366	3,710.0	40.000	3,484.4	240.05	75.037	98	0.94	037.42	0
037	Zinc (mg / kg (ppm))	0035	3,724.5	57.000	3,484.4	240.05	75.037	98	1.00	037.99	0
037	Zinc (mg / kg (ppm))	0619	3,729.0	68.000	3,484.4	240.05	75.037	98	1.02	037.41	0
037	Zinc (mg / kg (ppm))	0726	3,745.0	2.0000	3,484.4	240.05	75.037	98	1.09	037.42	0
037	Zinc (mg / kg (ppm))	0003	3,755.5	137.00	3,484.4	240.05	75.037	98	1.13	037.41	0
037	Zinc (mg / kg (ppm))	0354	3,760.4	8.1600	3,484.4	240.05	75.037	98	1.15	037.31	0
037	Zinc (mg / kg (ppm))	0027	3,768.7	131.65	3,484.4	240.05	75.037	98	1.18	037.43	0
037	Zinc (mg / kg (ppm))	0910	3,776.5	89.000	3,484.4	240.05	75.037	98	1.22	037.41	0
037	Zinc (mg / kg (ppm))	0353	3,788.0	404.00	3,484.4	240.05	75.037	98	1.26	037.43	0
037	Zinc (mg / kg (ppm))	0100	3,788.0	258.00	3,484.4	240.05	75.037	98	1.26	037.99	0
037	Zinc (mg / kg (ppm))	0297	3,793.5	417.00	3,484.4	240.05	75.037	98	1.29	037.43	0
037	Zinc (mg / kg (ppm))	0226	3,824.5	37.000	3,484.4	240.05	75.037	98	1.42	037.41	0
037	Zinc (mg / kg (ppm))	0675	3,852.0	43.170	3,484.4	240.05	75.037	98	1.53	037.43	0
037	Zinc (mg / kg (ppm))	0968	3,864.0	18.000	3,484.4	240.05	75.037	98	1.58	037.43	0
037	Zinc (mg / kg (ppm))	0035	3,890.5	7.0000	3,484.4	240.05	75.037	98	1.69	037.43	0
037	Zinc (mg / kg (ppm))	0553	3,915.0	30.000	3,484.4	240.05	75.037	98	1.79	037.53	0
037	Zinc (mg / kg (ppm))	2053	3,955.0	258.00	3,484.4	240.05	75.037	98	1.96	037.42	0
037	Zinc (mg / kg (ppm))	0035	4,013.0	8.0000	3,484.4	240.05	75.037	98	2.20	037.42	0
037	Zinc (mg / kg (ppm))	0042	3,733.0	756.00	3,484.4	240.05	75.037	98	1.04	037.42	1
037	Zinc (mg / kg (ppm))	0692	5,020.0	760.00	3,484.4	240.05	75.037	98	6.40	037.42	1
037	Zinc (mg / kg (ppm))	0169	0.00000	0.00000	3,484.4	240.05	75.037	98		037.43	4
038	Molybdenum (mg / kg (ppm))	0171	0.90000	0.20000	1.5801	0.23593	0.12855	15	-2.88	038.41	0
038	Molybdenum (mg / kg (ppm))	0353	1.3815	0.01860	1.5801	0.23593	0.12855	15	-0.84	038.43	0
038	Molybdenum (mg / kg (ppm))	0345	1.3850	0.13000	1.5801	0.23593	0.12855	15	-0.83	038.43	0
038	Molybdenum (mg / kg (ppm))	0098	1.4190	0.19200	1.5801	0.23593	0.12855	15	-0.68	038.53	0
038	Molybdenum (mg / kg (ppm))	0186	1.4400	0.08000	1.5801	0.23593	0.12855	15	-0.59	038.52	0
038	Molybdenum (mg / kg (ppm))	2055	1.4600	0.22000	1.5801	0.23593	0.12855	15	-0.51	038.53	0
038	Molybdenum (mg / kg (ppm))	0226	1.4650	0.03000	1.5801	0.23593	0.12855	15	-0.49	038.41	0
038	Molybdenum (mg / kg (ppm))	0910	1.5450	0.01000	1.5801	0.23593	0.12855	15	-0.15	038.52	0
038	Molybdenum (mg / kg (ppm))	0693	1.6112	0.07780	1.5801	0.23593	0.12855	15	0.13	038.42	0
038	Molybdenum (mg / kg (ppm))	0918	1.6602	0.11980	1.5801	0.23593	0.12855	15	0.34	038.53	0
038	Molybdenum (mg / kg (ppm))	0510	1.7000	0.00000	1.5801	0.23593	0.12855	15	0.51	038.43	0
038	Molybdenum (mg / kg (ppm))	0553	1.8050	0.07000	1.5801	0.23593	0.12855	15	0.95	038.53	0
038	Molybdenum (mg / kg (ppm))	2053	1.8100	0.02000	1.5801	0.23593	0.12855	15	0.97	038.42	0
038	Molybdenum (mg / kg (ppm))	0297	1.8600	0.18000	1.5801	0.23593	0.12855	15	1.19	038.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
038	Molybdenum (mg / kg (ppm))	0278	2.3700	0.58000	1.5801	0.23593	0.12855	15	3.35	038.42	0
038	Molybdenum (mg / kg (ppm))	0045	1.9750	0.93000	1.5801	0.23593	0.12855	15	1.67	038.42	1
038	Molybdenum (mg / kg (ppm))	0169	0.00000	0.00000	1.5801	0.23593	0.12855	15		038.43	4
040	Barium (mg / kg (ppm))	0918	9.6267	0.93870				2		040.53	0
040	Barium (mg / kg (ppm))	0560	9.9600	0.96000				2		040.42	0
041	Vanadium (mg / kg (ppm))	0553	1.2350	0.03000				1		041.53	0
042	Chloride (%)	0948	0.86000	0.08000				3		042.00	0
042	Chloride (%)	0413	0.94500	0.03000				3		042.02	0
042	Chloride (%)	0297	1.0700	0.16000				3		042.00	0
101	Choline Chloride (mg / kg (ppm))	0644	1,448.0	16.000	2,282.7	669.84	130.58	4	-1.25	101.02	0
101	Choline Chloride (mg / kg (ppm))	0675	2,042.9	166.30	2,282.7	669.84	130.58	4	-0.36	101.00	0
101	Choline Chloride (mg / kg (ppm))	2004	2,735.0	210.00	2,282.7	669.84	130.58	4	0.68	101.01	0
101	Choline Chloride (mg / kg (ppm))	0227	2,905.0	130.00	2,282.7	669.84	130.58	4	0.93	101.00	0
102	Niacin (mg / kg (ppm))	0227	112.00	6.0000				2		102.01	0
102	Niacin (mg / kg (ppm))	0675	152.79	5.8100				2		102.02	0
103	Pantothenic Acid (mg / kg (ppm))	0227	55.750	2.5000				2		103.01	0
103	Pantothenic Acid (mg / kg (ppm))	0629	68.950	0.50000				2		103.02	0
104	Riboflavin (mg / kg (ppm))	0910	16.910	3.1200	18.965	2.0693	1.3050	4	-0.99	104.03	0
104	Riboflavin (mg / kg (ppm))	0227	17.950	0.50000	18.965	2.0693	1.3050	4	-0.49	104.00	0
104	Riboflavin (mg / kg (ppm))	0208	19.300	1.0000	18.965	2.0693	1.3050	4	0.16	104.03	0
104	Riboflavin (mg / kg (ppm))	0171	21.700	0.60000	18.965	2.0693	1.3050	4	1.32	104.00	0
105	Thiamine (mg / kg (ppm))	0629	5.0500	0.00000	6.9414	1.8473	0.41075	4	-1.02	105.00	0
105	Thiamine (mg / kg (ppm))	0208	6.2055	0.98300	6.9414	1.8473	0.41075	4	-0.40	105.00	0
105	Thiamine (mg / kg (ppm))	0910	7.1000	0.22000	6.9414	1.8473	0.41075	4	0.09	105.00	0
105	Thiamine (mg / kg (ppm))	0227	9.4100	0.44000	6.9414	1.8473	0.41075	4	1.34	105.01	0
106	Vitamin A (KU / kg)	0027	10.392	0.12600	37.897	15.720	4.4437	23	-1.75	106.02	0
106	Vitamin A (KU / kg)	0675	17.470	1.2400	37.897	15.720	4.4437	23	-1.30	106.02	0
106	Vitamin A (KU / kg)	0619	20.800	0.00000	37.897	15.720	4.4437	23	-1.09	106.02	0
106	Vitamin A (KU / kg)	0028	22.867	0.53300	37.897	15.720	4.4437	23	-0.96	106.02	0
106	Vitamin A (KU / kg)	2068	23.000	0.00000	37.897	15.720	4.4437	23	-0.95	106.02	0
106	Vitamin A (KU / kg)	2066	25.650	1.1000	37.897	15.720	4.4437	23	-0.78	106.02	0
106	Vitamin A (KU / kg)	2053	26.350	6.5000	37.897	15.720	4.4437	23	-0.73	106.02	0
106	Vitamin A (KU / kg)	0676	29.805	10.030	37.897	15.720	4.4437	23	-0.51	106.02	0
106	Vitamin A (KU / kg)	0208	32.849	3.5270	37.897	15.720	4.4437	23	-0.32	106.02	0
106	Vitamin A (KU / kg)	0227	38.100	1.6000	37.897	15.720	4.4437	23	0.01	106.02	0
106	Vitamin A (KU / kg)	0610	38.650	5.1000	37.897	15.720	4.4437	23	0.05	106.02	0
106	Vitamin A (KU / kg)	0148	39.245	1.6700	37.897	15.720	4.4437	23	0.09	106.02	0
106	Vitamin A (KU / kg)	0910	39.700	1.2000	37.897	15.720	4.4437	23	0.11	106.02	0
106	Vitamin A (KU / kg)	2004	41.000	2.0000	37.897	15.720	4.4437	23	0.20	106.02	0
106	Vitamin A (KU / kg)	0941	42.337	0.00000	37.897	15.720	4.4437	23	0.28	106.02	0
106	Vitamin A (KU / kg)	0004	44.700	5.6400	37.897	15.720	4.4437	23	0.43	106.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
106	Vitamin A (KU / kg)	2103	45.170	6.1200	37.897	15.720	4.4437	23	0.46	106.02	0
106	Vitamin A (KU / kg)	0171	47.300	2.4000	37.897	15.720	4.4437	23	0.60	106.00	0
106	Vitamin A (KU / kg)	0563	47.400	0.44000	37.897	15.720	4.4437	23	0.60	106.02	0
106	Vitamin A (KU / kg)	0169	50.500	1.0000	37.897	15.720	4.4437	23	0.80	106.02	0
106	Vitamin A (KU / kg)	0098	67.000	15.200	37.897	15.720	4.4437	23	1.85	106.01	0
106	Vitamin A (KU / kg)	0861	70.975	35.530	37.897	15.720	4.4437	23	2.10	106.02	0
106	Vitamin A (KU / kg)	0905	79.615	1.2500	37.897	15.720	4.4437	23	2.65	106.02	0
106	Vitamin A (KU / kg)	0019	20.955	811.00	37.897	15.720	4.4437	23	1330.61	106.00	2
107	Vitamin B12 (µg / kg (ppb))	0227	93.650	3.3000				1		107.00	0
108	Vitamin D3 (KU / kg)	0675	5.0500	0.16000	8.7588	3.0719	1.9975	8	-1.21	108.02	0
108	Vitamin D3 (KU / kg)	0169	6.5000	1.0000	8.7588	3.0719	1.9975	8	-0.74	108.01	0
108	Vitamin D3 (KU / kg)	0610	7.4500	0.10000	8.7588	3.0719	1.9975	8	-0.43	108.02	0
108	Vitamin D3 (KU / kg)	0861	7.7600	3.0400	8.7588	3.0719	1.9975	8	-0.33	108.02	0
108	Vitamin D3 (KU / kg)	0208	8.0450	4.7100	8.7588	3.0719	1.9975	8	-0.23	108.02	0
108	Vitamin D3 (KU / kg)	0227	10.755	5.6900	8.7588	3.0719	1.9975	8	0.65	108.02	0
108	Vitamin D3 (KU / kg)	2004	11.600	1.2000	8.7588	3.0719	1.9975	8	0.92	108.99	0
108	Vitamin D3 (KU / kg)	0676	12.910	0.08000	8.7588	3.0719	1.9975	8	1.35	108.02	0
109	Vitamin E (IU/kg)	0619	0.11900	0.00000	187.14	22.895	7.1810	21	-8.17	109.02	0
109	Vitamin E (IU/kg)	2068	134.00	6.0000	187.14	22.895	7.1810	21	-2.32	109.02	0
109	Vitamin E (IU/kg)	0171	158.50	1.0000	187.14	22.895	7.1810	21	-1.25	109.99	0
109	Vitamin E (IU/kg)	0208	171.28	1.8220	187.14	22.895	7.1810	21	-0.69	109.02	0
109	Vitamin E (IU/kg)	0590	172.00	0.00000	187.14	22.895	7.1810	21	-0.66	109.02	0
109	Vitamin E (IU/kg)	0027	173.94	6.3100	187.14	22.895	7.1810	21	-0.58	109.02	0
109	Vitamin E (IU/kg)	2103	178.50	1.0000	187.14	22.895	7.1810	21	-0.38	109.02	0
109	Vitamin E (IU/kg)	0169	180.50	9.0000	187.14	22.895	7.1810	21	-0.29	109.02	0
109	Vitamin E (IU/kg)	0098	182.50	31.000	187.14	22.895	7.1810	21	-0.20	109.02	0
109	Vitamin E (IU/kg)	0905	183.50	1.0000	187.14	22.895	7.1810	21	-0.16	109.02	0
109	Vitamin E (IU/kg)	0610	185.25	23.500	187.14	22.895	7.1810	21	-0.08	109.02	0
109	Vitamin E (IU/kg)	0941	190.04	0.00000	187.14	22.895	7.1810	21	0.13	109.02	0
109	Vitamin E (IU/kg)	0910	190.90	3.6000	187.14	22.895	7.1810	21	0.16	109.02	0
109	Vitamin E (IU/kg)	0676	192.28	1.9900	187.14	22.895	7.1810	21	0.22	109.02	0
109	Vitamin E (IU/kg)	0955	194.50	15.000	187.14	22.895	7.1810	21	0.32	109.02	0
109	Vitamin E (IU/kg)	0861	200.00	0.00000	187.14	22.895	7.1810	21	0.56	109.02	0
109	Vitamin E (IU/kg)	2022	205.00	30.000	187.14	22.895	7.1810	21	0.78	109.02	0
109	Vitamin E (IU/kg)	0148	206.63	5.6000	187.14	22.895	7.1810	21	0.85	109.02	0
109	Vitamin E (IU/kg)	0227	216.06	7.8000	187.14	22.895	7.1810	21	1.26	109.02	0
109	Vitamin E (IU/kg)	0675	241.26	3.1800	187.14	22.895	7.1810	21	2.36	109.02	0
109	Vitamin E (IU/kg)	0563	246.50	3.0000	187.14	22.895	7.1810	21	2.59	109.02	0
109	Vitamin E (IU/kg)	2066	3,975.0	170.00	187.14	22.895	7.1810	21	165.45	109.02	2
112	Pyridoxine (µg / g)	0227	7.1850	0.03000				1		112.01	0
113	Folic Acid (mg / kg (ppm))	0227	3.4900	0.26000				1		113.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
114	Biotin (mg / kg (ppm))	0227	1.7850	0.13000				1		114.01	0
120	Alanine (%)	0626	0.95300	0.05200	1.0314	0.03831	0.02070	26	-2.05	120.05	0
120	Alanine (%)	0148	0.96500	0.01000	1.0314	0.03831	0.02070	26	-1.73	120.05	0
120	Alanine (%)	0610	0.97500	0.01000	1.0314	0.03831	0.02070	26	-1.47	120.05	0
120	Alanine (%)	0941	0.98000	0.04000	1.0314	0.03831	0.02070	26	-1.34	120.00	0
120	Alanine (%)	0227	0.99000	0.00000	1.0314	0.03831	0.02070	26	-1.08	120.00	0
120	Alanine (%)	0676	1.0065	0.06700	1.0314	0.03831	0.02070	26	-0.65	120.05	0
120	Alanine (%)	0354	1.0185	0.00500	1.0314	0.03831	0.02070	26	-0.34	120.00	0
120	Alanine (%)	0684	1.0190	0.03000	1.0314	0.03831	0.02070	26	-0.32	120.00	0
120	Alanine (%)	0652	1.0200	0.02000	1.0314	0.03831	0.02070	26	-0.30	120.00	0
120	Alanine (%)	0872	1.0250	0.00000	1.0314	0.03831	0.02070	26	-0.17	120.00	0
120	Alanine (%)	0504	1.0300	0.04000	1.0314	0.03831	0.02070	26	-0.04	120.00	0
120	Alanine (%)	0644	1.0305	0.00500	1.0314	0.03831	0.02070	26	-0.02	120.00	0
120	Alanine (%)	2059	1.0345	0.00300	1.0314	0.03831	0.02070	26	0.08	120.00	0
120	Alanine (%)	0619	1.0350	0.01000	1.0314	0.03831	0.02070	26	0.09	120.00	0
120	Alanine (%)	0878	1.0355	0.01100	1.0314	0.03831	0.02070	26	0.11	120.00	0
120	Alanine (%)	0870	1.0362	0.04870	1.0314	0.03831	0.02070	26	0.12	120.00	0
120	Alanine (%)	0859	1.0395	0.01500	1.0314	0.03831	0.02070	26	0.21	120.00	0
120	Alanine (%)	0571	1.0470	0.00200	1.0314	0.03831	0.02070	26	0.41	120.00	0
120	Alanine (%)	0675	1.0500	0.00000	1.0314	0.03831	0.02070	26	0.48	120.00	0
120	Alanine (%)	0910	1.0500	0.06000	1.0314	0.03831	0.02070	26	0.48	120.00	0
120	Alanine (%)	0098	1.0555	0.00500	1.0314	0.03831	0.02070	26	0.63	120.02	0
120	Alanine (%)	0968	1.0625	0.02500	1.0314	0.03831	0.02070	26	0.81	120.00	0
120	Alanine (%)	0171	1.0675	0.00700	1.0314	0.03831	0.02070	26	0.94	120.00	0
120	Alanine (%)	0918	1.0840	0.02240	1.0314	0.03831	0.02070	26	1.37	120.00	0
120	Alanine (%)	0904	1.1100	0.04000	1.0314	0.03831	0.02070	26	2.05	120.99	0
120	Alanine (%)	0955	1.1450	0.01000	1.0314	0.03831	0.02070	26	2.96	120.00	0
121	Arginine (%)	0870	1.1660	0.05490	1.2748	0.04602	0.02448	25	-2.37	121.00	0
121	Arginine (%)	0626	1.2040	0.01000	1.2748	0.04602	0.02448	25	-1.54	121.05	0
121	Arginine (%)	0941	1.2100	0.02000	1.2748	0.04602	0.02448	25	-1.41	121.00	0
121	Arginine (%)	0918	1.2414	0.01610	1.2748	0.04602	0.02448	25	-0.73	121.00	0
121	Arginine (%)	0684	1.2415	0.04700	1.2748	0.04602	0.02448	25	-0.72	121.00	0
121	Arginine (%)	0859	1.2455	0.00900	1.2748	0.04602	0.02448	25	-0.64	121.00	0
121	Arginine (%)	0354	1.2520	0.00800	1.2748	0.04602	0.02448	25	-0.50	121.00	0
121	Arginine (%)	0227	1.2550	0.01000	1.2748	0.04602	0.02448	25	-0.43	121.00	0
121	Arginine (%)	0171	1.2560	0.00800	1.2748	0.04602	0.02448	25	-0.41	121.00	0
121	Arginine (%)	0910	1.2600	0.06000	1.2748	0.04602	0.02448	25	-0.32	121.00	0
121	Arginine (%)	0878	1.2640	0.02000	1.2748	0.04602	0.02448	25	-0.23	121.00	0
121	Arginine (%)	0504	1.2650	0.03000	1.2748	0.04602	0.02448	25	-0.21	121.00	0
121	Arginine (%)	0619	1.2700	0.02000	1.2748	0.04602	0.02448	25	-0.10	121.00	0
121	Arginine (%)	0644	1.2785	0.01500	1.2748	0.04602	0.02448	25	0.08	121.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
121	Arginine (%)	0968	1.2790	0.01200	1.2748	0.04602	0.02448	25	0.09	121.00	0
121	Arginine (%)	2059	1.2800	0.00800	1.2748	0.04602	0.02448	25	0.11	121.00	0
121	Arginine (%)	0652	1.2900	0.00000	1.2748	0.04602	0.02448	25	0.33	121.00	0
121	Arginine (%)	0676	1.2905	0.09300	1.2748	0.04602	0.02448	25	0.34	121.05	0
121	Arginine (%)	0098	1.3035	0.01100	1.2748	0.04602	0.02448	25	0.62	121.02	0
121	Arginine (%)	0610	1.3050	0.03000	1.2748	0.04602	0.02448	25	0.66	121.05	0
121	Arginine (%)	0872	1.3070	0.00400	1.2748	0.04602	0.02448	25	0.70	121.00	0
121	Arginine (%)	0571	1.3330	0.02600	1.2748	0.04602	0.02448	25	1.26	121.00	0
121	Arginine (%)	0904	1.3500	0.04000	1.2748	0.04602	0.02448	25	1.63	121.99	0
121	Arginine (%)	0955	1.3550	0.05000	1.2748	0.04602	0.02448	25	1.74	121.00	0
121	Arginine (%)	0675	1.4550	0.01000	1.2748	0.04602	0.02448	25	3.92	121.00	0
121	Arginine (%)	0148	1.1400	0.22000	1.2748	0.04602	0.02448	25	-2.93	121.05	1
122	Aspartic (%)	0941	1.8700	0.02000	2.0150	0.05416	0.02730	25	-2.68	122.00	0
122	Aspartic (%)	0626	1.9210	0.03800	2.0150	0.05416	0.02730	25	-1.74	122.05	0
122	Aspartic (%)	0504	1.9450	0.05000	2.0150	0.05416	0.02730	25	-1.29	122.00	0
122	Aspartic (%)	0354	1.9555	0.01300	2.0150	0.05416	0.02730	25	-1.10	122.00	0
122	Aspartic (%)	0148	1.9650	0.11000	2.0150	0.05416	0.02730	25	-0.92	122.05	0
122	Aspartic (%)	0684	1.9800	0.06800	2.0150	0.05416	0.02730	25	-0.65	122.00	0
122	Aspartic (%)	0644	1.9845	0.01100	2.0150	0.05416	0.02730	25	-0.56	122.00	0
122	Aspartic (%)	0910	1.9950	0.01000	2.0150	0.05416	0.02730	25	-0.37	122.00	0
122	Aspartic (%)	0904	1.9950	0.01000	2.0150	0.05416	0.02730	25	-0.37	122.99	0
122	Aspartic (%)	0859	2.0025	0.00900	2.0150	0.05416	0.02730	25	-0.23	122.00	0
122	Aspartic (%)	0610	2.0050	0.05000	2.0150	0.05416	0.02730	25	-0.18	122.05	0
122	Aspartic (%)	0872	2.0070	0.00200	2.0150	0.05416	0.02730	25	-0.15	122.00	0
122	Aspartic (%)	0675	2.0100	0.02000	2.0150	0.05416	0.02730	25	-0.09	122.00	0
122	Aspartic (%)	2059	2.0165	0.01100	2.0150	0.05416	0.02730	25	0.03	122.00	0
122	Aspartic (%)	0870	2.0248	0.07270	2.0150	0.05416	0.02730	25	0.18	122.00	0
122	Aspartic (%)	0878	2.0320	0.02000	2.0150	0.05416	0.02730	25	0.31	122.00	0
122	Aspartic (%)	0227	2.0350	0.01000	2.0150	0.05416	0.02730	25	0.37	122.00	0
122	Aspartic (%)	0571	2.0405	0.00500	2.0150	0.05416	0.02730	25	0.47	122.00	0
122	Aspartic (%)	0652	2.0450	0.03000	2.0150	0.05416	0.02730	25	0.55	122.00	0
122	Aspartic (%)	0968	2.0660	0.01800	2.0150	0.05416	0.02730	25	0.94	122.00	0
122	Aspartic (%)	0098	2.0675	0.00300	2.0150	0.05416	0.02730	25	0.97	122.02	0
122	Aspartic (%)	0171	2.0685	0.00500	2.0150	0.05416	0.02730	25	0.99	122.00	0
122	Aspartic (%)	0619	2.0750	0.01000	2.0150	0.05416	0.02730	25	1.11	122.00	0
122	Aspartic (%)	0955	2.1150	0.03000	2.0150	0.05416	0.02730	25	1.85	122.00	0
122	Aspartic (%)	0918	2.1271	0.05690	2.0150	0.05416	0.02730	25	2.07	122.00	0
122	Aspartic (%)	0676	2.0195	0.16500	2.0150	0.05416	0.02730	25	0.08	122.05	1
124	Cysteine/Cystine (%)	0870	0.25675	0.00770	0.31064	0.01758	0.00566	24	-3.07	124.00	0
124	Cysteine/Cystine (%)	0684	0.27500	0.00200	0.31064	0.01758	0.00566	24	-2.03	124.00	0
124	Cysteine/Cystine (%)	0227	0.28000	0.00000	0.31064	0.01758	0.00566	24	-1.74	124.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
124	Cysteine/Cystine (%)	0904	0.29000	0.02000	0.31064	0.01758	0.00566	24	-1.17	124.99	0
124	Cysteine/Cystine (%)	0955	0.30150	0.00700	0.31064	0.01758	0.00566	24	-0.52	124.00	0
124	Cysteine/Cystine (%)	2059	0.30350	0.00700	0.31064	0.01758	0.00566	24	-0.41	124.00	0
124	Cysteine/Cystine (%)	0652	0.30500	0.01000	0.31064	0.01758	0.00566	24	-0.32	124.00	0
124	Cysteine/Cystine (%)	0872	0.30600	0.00400	0.31064	0.01758	0.00566	24	-0.26	124.00	0
124	Cysteine/Cystine (%)	0878	0.30600	0.00000	0.31064	0.01758	0.00566	24	-0.26	124.00	0
124	Cysteine/Cystine (%)	0968	0.30700	0.00200	0.31064	0.01758	0.00566	24	-0.21	124.00	0
124	Cysteine/Cystine (%)	0504	0.31000	0.00000	0.31064	0.01758	0.00566	24	-0.04	124.00	0
124	Cysteine/Cystine (%)	0644	0.31000	0.00200	0.31064	0.01758	0.00566	24	-0.04	124.00	0
124	Cysteine/Cystine (%)	0910	0.31000	0.02000	0.31064	0.01758	0.00566	24	-0.04	124.00	0
124	Cysteine/Cystine (%)	0941	0.31000	0.00000	0.31064	0.01758	0.00566	24	-0.04	124.00	0
124	Cysteine/Cystine (%)	0676	0.31000	0.00000	0.31064	0.01758	0.00566	24	-0.04	124.05	0
124	Cysteine/Cystine (%)	0918	0.31250	0.00420	0.31064	0.01758	0.00566	24	0.11	124.00	0
124	Cysteine/Cystine (%)	0619	0.31300	0.00400	0.31064	0.01758	0.00566	24	0.13	124.00	0
124	Cysteine/Cystine (%)	0354	0.32150	0.00100	0.31064	0.01758	0.00566	24	0.62	124.00	0
124	Cysteine/Cystine (%)	0571	0.32250	0.01300	0.31064	0.01758	0.00566	24	0.67	124.00	0
124	Cysteine/Cystine (%)	0098	0.32800	0.00400	0.31064	0.01758	0.00566	24	0.99	124.02	0
124	Cysteine/Cystine (%)	0675	0.33000	0.02000	0.31064	0.01758	0.00566	24	1.10	124.00	0
124	Cysteine/Cystine (%)	0859	0.33200	0.00200	0.31064	0.01758	0.00566	24	1.22	124.00	0
124	Cysteine/Cystine (%)	0171	0.33700	0.00600	0.31064	0.01758	0.00566	24	1.50	124.00	0
124	Cysteine/Cystine (%)	0610	0.39000	0.00000	0.31064	0.01758	0.00566	24	4.51	124.05	0
124	Cysteine/Cystine (%)	0148	0.36500	0.05000	0.31064	0.01758	0.00566	24	3.09	124.05	1
125	Glutamic (%)	0941	3.1250	0.03000	3.3613	0.10203	0.05955	26	-2.32	125.00	0
125	Glutamic (%)	0904	3.2150	0.15000	3.3613	0.10203	0.05955	26	-1.43	125.99	0
125	Glutamic (%)	0626	3.2355	0.08500	3.3613	0.10203	0.05955	26	-1.23	125.05	0
125	Glutamic (%)	0504	3.2550	0.07000	3.3613	0.10203	0.05955	26	-1.04	125.00	0
125	Glutamic (%)	0354	3.2610	0.02800	3.3613	0.10203	0.05955	26	-0.98	125.00	0
125	Glutamic (%)	0872	3.2635	0.01500	3.3613	0.10203	0.05955	26	-0.96	125.00	0
125	Glutamic (%)	0684	3.2940	0.13000	3.3613	0.10203	0.05955	26	-0.66	125.00	0
125	Glutamic (%)	0148	3.3150	0.11000	3.3613	0.10203	0.05955	26	-0.45	125.05	0
125	Glutamic (%)	0644	3.3195	0.01100	3.3613	0.10203	0.05955	26	-0.41	125.00	0
125	Glutamic (%)	0859	3.3200	0.01600	3.3613	0.10203	0.05955	26	-0.40	125.00	0
125	Glutamic (%)	0610	3.3350	0.07000	3.3613	0.10203	0.05955	26	-0.26	125.05	0
125	Glutamic (%)	0227	3.3550	0.01000	3.3613	0.10203	0.05955	26	-0.06	125.00	0
125	Glutamic (%)	2059	3.3655	0.00500	3.3613	0.10203	0.05955	26	0.04	125.00	0
125	Glutamic (%)	0878	3.3820	0.06200	3.3613	0.10203	0.05955	26	0.20	125.00	0
125	Glutamic (%)	0676	3.3845	0.20500	3.3613	0.10203	0.05955	26	0.23	125.05	0
125	Glutamic (%)	0571	3.3870	0.06200	3.3613	0.10203	0.05955	26	0.25	125.00	0
125	Glutamic (%)	0910	3.3900	0.08000	3.3613	0.10203	0.05955	26	0.28	125.00	0
125	Glutamic (%)	0171	3.4010	0.00000	3.3613	0.10203	0.05955	26	0.39	125.00	0
125	Glutamic (%)	0675	3.4150	0.01000	3.3613	0.10203	0.05955	26	0.53	125.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
125	Glutamic (%)	0870	3.4247	0.13550	3.3613	0.10203	0.05955	26	0.62	125.00	0
125	Glutamic (%)	0968	3.4300	0.03000	3.3613	0.10203	0.05955	26	0.67	125.00	0
125	Glutamic (%)	0098	3.4395	0.00900	3.3613	0.10203	0.05955	26	0.77	125.02	0
125	Glutamic (%)	0918	3.4646	0.01480	3.3613	0.10203	0.05955	26	1.01	125.00	0
125	Glutamic (%)	0652	3.5050	0.05000	3.3613	0.10203	0.05955	26	1.41	125.00	0
125	Glutamic (%)	0619	3.6600	0.02000	3.3613	0.10203	0.05955	26	2.93	125.00	0
125	Glutamic (%)	0955	3.8400	0.14000	3.3613	0.10203	0.05955	26	4.69	125.00	0
126	Glycine (%)	0148	0.93000	0.06000	0.98964	0.02400	0.01646	25	-2.49	126.05	0
126	Glycine (%)	0941	0.94500	0.03000	0.98964	0.02400	0.01646	25	-1.86	126.00	0
126	Glycine (%)	0171	0.95000	0.00400	0.98964	0.02400	0.01646	25	-1.65	126.00	0
126	Glycine (%)	0684	0.96250	0.00300	0.98964	0.02400	0.01646	25	-1.13	126.00	0
126	Glycine (%)	0354	0.97000	0.00200	0.98964	0.02400	0.01646	25	-0.82	126.00	0
126	Glycine (%)	0872	0.97300	0.00000	0.98964	0.02400	0.01646	25	-0.69	126.00	0
126	Glycine (%)	0910	0.97500	0.01000	0.98964	0.02400	0.01646	25	-0.61	126.00	0
126	Glycine (%)	0652	0.98000	0.02000	0.98964	0.02400	0.01646	25	-0.40	126.00	0
126	Glycine (%)	0904	0.98000	0.04000	0.98964	0.02400	0.01646	25	-0.40	126.99	0
126	Glycine (%)	0626	0.98850	0.00300	0.98964	0.02400	0.01646	25	-0.05	126.05	0
126	Glycine (%)	0859	0.99000	0.00200	0.98964	0.02400	0.01646	25	0.01	126.00	0
126	Glycine (%)	0098	0.99000	0.00600	0.98964	0.02400	0.01646	25	0.01	126.02	0
126	Glycine (%)	0610	0.99000	0.02000	0.98964	0.02400	0.01646	25	0.01	126.05	0
126	Glycine (%)	0227	0.99500	0.01000	0.98964	0.02400	0.01646	25	0.22	126.00	0
126	Glycine (%)	0878	0.99600	0.02000	0.98964	0.02400	0.01646	25	0.26	126.00	0
126	Glycine (%)	0870	0.99720	0.03700	0.98964	0.02400	0.01646	25	0.31	126.00	0
126	Glycine (%)	2059	0.99800	0.00600	0.98964	0.02400	0.01646	25	0.35	126.00	0
126	Glycine (%)	0968	0.99900	0.00400	0.98964	0.02400	0.01646	25	0.39	126.00	0
126	Glycine (%)	0644	1.0040	0.01000	0.98964	0.02400	0.01646	25	0.60	126.00	0
126	Glycine (%)	0571	1.0050	0.00600	0.98964	0.02400	0.01646	25	0.64	126.00	0
126	Glycine (%)	0675	1.0050	0.01000	0.98964	0.02400	0.01646	25	0.64	126.00	0
126	Glycine (%)	0918	1.0057	0.02850	0.98964	0.02400	0.01646	25	0.67	126.00	0
126	Glycine (%)	0504	1.0250	0.07000	0.98964	0.02400	0.01646	25	1.47	126.00	0
126	Glycine (%)	0619	1.0700	0.00000	0.98964	0.02400	0.01646	25	3.35	126.00	0
126	Glycine (%)	0955	1.0950	0.01000	0.98964	0.02400	0.01646	25	4.39	126.00	0
126	Glycine (%)	0676	1.0530	0.09200	0.98964	0.02400	0.01646	25	2.64	126.05	1
127	Histidine (%)	0859	0.46200	0.00600	0.49121	0.01750	0.01034	24	-1.67	127.00	0
127	Histidine (%)	0870	0.46770	0.01520	0.49121	0.01750	0.01034	24	-1.34	127.00	0
127	Histidine (%)	0684	0.46950	0.05100	0.49121	0.01750	0.01034	24	-1.24	127.00	0
127	Histidine (%)	0626	0.47300	0.00200	0.49121	0.01750	0.01034	24	-1.04	127.05	0
127	Histidine (%)	0354	0.47800	0.00400	0.49121	0.01750	0.01034	24	-0.75	127.00	0
127	Histidine (%)	0098	0.47950	0.00100	0.49121	0.01750	0.01034	24	-0.67	127.02	0
127	Histidine (%)	0941	0.48000	0.00000	0.49121	0.01750	0.01034	24	-0.64	127.00	0
127	Histidine (%)	0571	0.48450	0.00500	0.49121	0.01750	0.01034	24	-0.38	127.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
127	Histidine (%)	0504	0.48500	0.01000	0.49121	0.01750	0.01034	24	-0.35	127.00	0
127	Histidine (%)	0610	0.48500	0.01000	0.49121	0.01750	0.01034	24	-0.35	127.05	0
127	Histidine (%)	2059	0.48850	0.00500	0.49121	0.01750	0.01034	24	-0.15	127.00	0
127	Histidine (%)	0227	0.49000	0.00000	0.49121	0.01750	0.01034	24	-0.07	127.00	0
127	Histidine (%)	0968	0.49350	0.00500	0.49121	0.01750	0.01034	24	0.13	127.00	0
127	Histidine (%)	0652	0.49500	0.01000	0.49121	0.01750	0.01034	24	0.22	127.00	0
127	Histidine (%)	0675	0.49500	0.01000	0.49121	0.01750	0.01034	24	0.22	127.00	0
127	Histidine (%)	0910	0.49500	0.03000	0.49121	0.01750	0.01034	24	0.22	127.00	0
127	Histidine (%)	0644	0.49550	0.00300	0.49121	0.01750	0.01034	24	0.25	127.00	0
127	Histidine (%)	0676	0.50000	0.02600	0.49121	0.01750	0.01034	24	0.50	127.05	0
127	Histidine (%)	0872	0.50300	0.00400	0.49121	0.01750	0.01034	24	0.67	127.00	0
127	Histidine (%)	0918	0.50395	0.00690	0.49121	0.01750	0.01034	24	0.73	127.00	0
127	Histidine (%)	0904	0.51000	0.02000	0.49121	0.01750	0.01034	24	1.07	127.99	0
127	Histidine (%)	0878	0.53400	0.01600	0.49121	0.01750	0.01034	24	2.44	127.00	0
127	Histidine (%)	0619	0.58550	0.00100	0.49121	0.01750	0.01034	24	5.39	127.00	0
127	Histidine (%)	0171	0.59750	0.00700	0.49121	0.01750	0.01034	24	6.07	127.00	0
127	Histidine (%)	0148	0.42000	0.06000	0.49121	0.01750	0.01034	24	-4.07	127.05	1
128	Isoleucine (%)	0684	0.70350	0.07100	0.85660	0.05291	0.01637	26	-2.89	128.00	0
128	Isoleucine (%)	0148	0.70500	0.03000	0.85660	0.05291	0.01637	26	-2.87	128.05	0
128	Isoleucine (%)	0870	0.77520	0.00360	0.85660	0.05291	0.01637	26	-1.54	128.00	0
128	Isoleucine (%)	0171	0.79750	0.00100	0.85660	0.05291	0.01637	26	-1.12	128.00	0
128	Isoleucine (%)	0676	0.80350	0.06100	0.85660	0.05291	0.01637	26	-1.00	128.05	0
128	Isoleucine (%)	0941	0.82000	0.02000	0.85660	0.05291	0.01637	26	-0.69	128.00	0
128	Isoleucine (%)	0859	0.82050	0.00500	0.85660	0.05291	0.01637	26	-0.68	128.00	0
128	Isoleucine (%)	0918	0.82655	0.00010	0.85660	0.05291	0.01637	26	-0.57	128.00	0
128	Isoleucine (%)	0626	0.83500	0.00600	0.85660	0.05291	0.01637	26	-0.41	128.05	0
128	Isoleucine (%)	0354	0.84000	0.00000	0.85660	0.05291	0.01637	26	-0.31	128.00	0
128	Isoleucine (%)	0652	0.85000	0.00000	0.85660	0.05291	0.01637	26	-0.12	128.00	0
128	Isoleucine (%)	0878	0.85250	0.01300	0.85660	0.05291	0.01637	26	-0.08	128.00	0
128	Isoleucine (%)	2059	0.86250	0.00500	0.85660	0.05291	0.01637	26	0.11	128.00	0
128	Isoleucine (%)	0968	0.87200	0.03000	0.85660	0.05291	0.01637	26	0.29	128.00	0
128	Isoleucine (%)	0610	0.87500	0.03000	0.85660	0.05291	0.01637	26	0.35	128.05	0
128	Isoleucine (%)	0872	0.87550	0.02700	0.85660	0.05291	0.01637	26	0.36	128.00	0
128	Isoleucine (%)	0571	0.87800	0.01000	0.85660	0.05291	0.01637	26	0.40	128.00	0
128	Isoleucine (%)	0227	0.88000	0.02000	0.85660	0.05291	0.01637	26	0.44	128.00	0
128	Isoleucine (%)	0644	0.88400	0.00600	0.85660	0.05291	0.01637	26	0.52	128.00	0
128	Isoleucine (%)	0098	0.88500	0.00600	0.85660	0.05291	0.01637	26	0.54	128.02	0
128	Isoleucine (%)	0675	0.89500	0.01000	0.85660	0.05291	0.01637	26	0.73	128.00	0
128	Isoleucine (%)	0619	0.89550	0.02100	0.85660	0.05291	0.01637	26	0.74	128.00	0
128	Isoleucine (%)	0910	0.90500	0.03000	0.85660	0.05291	0.01637	26	0.91	128.00	0
128	Isoleucine (%)	0904	0.91500	0.01000	0.85660	0.05291	0.01637	26	1.10	128.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
128	Isoleucine (%)	0504	0.94500	0.01000	0.85660	0.05291	0.01637	26	1.67	128.00	0
128	Isoleucine (%)	0955	1.0300	0.00000	0.85660	0.05291	0.01637	26	3.28	128.00	0
129	Leucine (%)	0684	1.3720	0.03400	1.5504	0.05451	0.01539	25	-3.27	129.00	0
129	Leucine (%)	0941	1.4650	0.05000	1.5504	0.05451	0.01539	25	-1.57	129.00	0
129	Leucine (%)	0870	1.4719	0.03530	1.5504	0.05451	0.01539	25	-1.44	129.00	0
129	Leucine (%)	0148	1.4800	0.00000	1.5504	0.05451	0.01539	25	-1.29	129.05	0
129	Leucine (%)	0626	1.5070	0.01200	1.5504	0.05451	0.01539	25	-0.80	129.05	0
129	Leucine (%)	0878	1.5170	0.03200	1.5504	0.05451	0.01539	25	-0.61	129.00	0
129	Leucine (%)	0171	1.5205	0.00500	1.5504	0.05451	0.01539	25	-0.55	129.00	0
129	Leucine (%)	2059	1.5380	0.00400	1.5504	0.05451	0.01539	25	-0.23	129.00	0
129	Leucine (%)	0675	1.5400	0.02000	1.5504	0.05451	0.01539	25	-0.19	129.00	0
129	Leucine (%)	0354	1.5410	0.01000	1.5504	0.05451	0.01539	25	-0.17	129.00	0
129	Leucine (%)	0872	1.5410	0.01200	1.5504	0.05451	0.01539	25	-0.17	129.00	0
129	Leucine (%)	0644	1.5435	0.00500	1.5504	0.05451	0.01539	25	-0.13	129.00	0
129	Leucine (%)	0859	1.5490	0.00600	1.5504	0.05451	0.01539	25	-0.03	129.00	0
129	Leucine (%)	0652	1.5500	0.00000	1.5504	0.05451	0.01539	25	-0.01	129.00	0
129	Leucine (%)	0227	1.5550	0.01000	1.5504	0.05451	0.01539	25	0.08	129.00	0
129	Leucine (%)	0910	1.5700	0.02000	1.5504	0.05451	0.01539	25	0.36	129.00	0
129	Leucine (%)	0571	1.5710	0.01400	1.5504	0.05451	0.01539	25	0.38	129.00	0
129	Leucine (%)	0098	1.5735	0.00300	1.5504	0.05451	0.01539	25	0.42	129.02	0
129	Leucine (%)	0968	1.5815	0.00300	1.5504	0.05451	0.01539	25	0.57	129.00	0
129	Leucine (%)	0504	1.5900	0.02000	1.5504	0.05451	0.01539	25	0.73	129.00	0
129	Leucine (%)	0610	1.5950	0.03000	1.5504	0.05451	0.01539	25	0.82	129.05	0
129	Leucine (%)	0904	1.6100	0.02000	1.5504	0.05451	0.01539	25	1.09	129.99	0
129	Leucine (%)	0918	1.6203	0.00940	1.5504	0.05451	0.01539	25	1.28	129.00	0
129	Leucine (%)	0619	1.6250	0.01000	1.5504	0.05451	0.01539	25	1.37	129.00	0
129	Leucine (%)	0955	1.7200	0.02000	1.5504	0.05451	0.01539	25	3.11	129.00	0
129	Leucine (%)	0676	1.5290	0.08000	1.5504	0.05451	0.01539	25	-0.39	129.05	1
130	L-Lysine (%)	0848	1.1900	0.06000	1.4318	0.06009	0.02606	30	-4.02	130.00	0
130	L-Lysine (%)	2145	1.2150	0.03000	1.4318	0.06009	0.02606	30	-3.61	130.01	0
130	L-Lysine (%)	0619	1.3200	0.02000	1.4318	0.06009	0.02606	30	-1.86	130.00	0
130	L-Lysine (%)	2053	1.3450	0.01000	1.4318	0.06009	0.02606	30	-1.44	130.99	0
130	L-Lysine (%)	0148	1.3600	0.06000	1.4318	0.06009	0.02606	30	-1.19	130.05	0
130	L-Lysine (%)	0941	1.3850	0.03000	1.4318	0.06009	0.02606	30	-0.78	130.00	0
130	L-Lysine (%)	0878	1.4010	0.01600	1.4318	0.06009	0.02606	30	-0.51	130.00	0
130	L-Lysine (%)	0872	1.4115	0.00500	1.4318	0.06009	0.02606	30	-0.34	130.00	0
130	L-Lysine (%)	0652	1.4150	0.01000	1.4318	0.06009	0.02606	30	-0.28	130.00	0
130	L-Lysine (%)	0676	1.4160	0.07200	1.4318	0.06009	0.02606	30	-0.26	130.05	0
130	L-Lysine (%)	0160	1.4200	0.00000	1.4318	0.06009	0.02606	30	-0.20	130.00	0
130	L-Lysine (%)	0675	1.4200	0.02000	1.4318	0.06009	0.02606	30	-0.20	130.00	0
130	L-Lysine (%)	0859	1.4220	0.00200	1.4318	0.06009	0.02606	30	-0.16	130.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
130	L-Lysine (%)	0684	1.4235	0.06700	1.4318	0.06009	0.02606	30	-0.14	130.00	0
130	L-Lysine (%)	0354	1.4255	0.00300	1.4318	0.06009	0.02606	30	-0.10	130.00	0
130	L-Lysine (%)	0626	1.4285	0.02100	1.4318	0.06009	0.02606	30	-0.05	130.05	0
130	L-Lysine (%)	0910	1.4300	0.02000	1.4318	0.06009	0.02606	30	-0.03	130.00	0
130	L-Lysine (%)	0968	1.4330	0.01000	1.4318	0.06009	0.02606	30	0.02	130.00	0
130	L-Lysine (%)	0571	1.4480	0.00600	1.4318	0.06009	0.02606	30	0.27	130.00	0
130	L-Lysine (%)	2059	1.4560	0.00200	1.4318	0.06009	0.02606	30	0.40	130.00	0
130	L-Lysine (%)	0644	1.4600	0.00200	1.4318	0.06009	0.02606	30	0.47	130.00	0
130	L-Lysine (%)	0610	1.4600	0.02000	1.4318	0.06009	0.02606	30	0.47	130.05	0
130	L-Lysine (%)	0227	1.4650	0.07000	1.4318	0.06009	0.02606	30	0.55	130.00	0
130	L-Lysine (%)	0918	1.4676	0.00080	1.4318	0.06009	0.02606	30	0.60	130.00	0
130	L-Lysine (%)	0904	1.4750	0.09000	1.4318	0.06009	0.02606	30	0.72	130.99	0
130	L-Lysine (%)	0870	1.4874	0.02190	1.4318	0.06009	0.02606	30	0.92	130.00	0
130	L-Lysine (%)	0098	1.5145	0.02100	1.4318	0.06009	0.02606	30	1.38	130.02	0
130	L-Lysine (%)	0504	1.5150	0.07000	1.4318	0.06009	0.02606	30	1.38	130.00	0
130	L-Lysine (%)	0955	1.6050	0.01000	1.4318	0.06009	0.02606	30	2.88	130.00	0
130	L-Lysine (%)	0171	1.6590	0.01200	1.4318	0.06009	0.02606	30	3.78	130.00	0
130	L-Lysine (%)	0014	1.4000	0.16000	1.4318	0.06009	0.02606	30	-0.53	130.05	1
131	Methionine (%)	0148	0.32500	0.01000	0.38524	0.02371	0.00984	28	-2.54	131.05	0
131	Methionine (%)	2053	0.32750	0.03700	0.38524	0.02371	0.00984	28	-2.44	131.99	0
131	Methionine (%)	0014	0.34750	0.00700	0.38524	0.02371	0.00984	28	-1.59	131.05	0
131	Methionine (%)	0955	0.35600	0.00600	0.38524	0.02371	0.00984	28	-1.23	131.00	0
131	Methionine (%)	0904	0.36500	0.01000	0.38524	0.02371	0.00984	28	-0.85	131.99	0
131	Methionine (%)	0941	0.37000	0.00000	0.38524	0.02371	0.00984	28	-0.64	131.00	0
131	Methionine (%)	0626	0.37050	0.03900	0.38524	0.02371	0.00984	28	-0.62	131.05	0
131	Methionine (%)	0171	0.37100	0.00800	0.38524	0.02371	0.00984	28	-0.60	131.00	0
131	Methionine (%)	0859	0.37100	0.00800	0.38524	0.02371	0.00984	28	-0.60	131.00	0
131	Methionine (%)	0098	0.37300	0.00400	0.38524	0.02371	0.00984	28	-0.52	131.02	0
131	Methionine (%)	0504	0.37500	0.01000	0.38524	0.02371	0.00984	28	-0.43	131.00	0
131	Methionine (%)	2059	0.38200	0.00400	0.38524	0.02371	0.00984	28	-0.14	131.00	0
131	Methionine (%)	0918	0.38275	0.00070	0.38524	0.02371	0.00984	28	-0.11	131.00	0
131	Methionine (%)	0684	0.38650	0.01700	0.38524	0.02371	0.00984	28	0.05	131.00	0
131	Methionine (%)	0878	0.38750	0.00500	0.38524	0.02371	0.00984	28	0.10	131.00	0
131	Methionine (%)	0354	0.39050	0.01300	0.38524	0.02371	0.00984	28	0.22	131.00	0
131	Methionine (%)	0872	0.39050	0.00700	0.38524	0.02371	0.00984	28	0.22	131.00	0
131	Methionine (%)	0968	0.39100	0.00200	0.38524	0.02371	0.00984	28	0.24	131.00	0
131	Methionine (%)	0571	0.39650	0.00500	0.38524	0.02371	0.00984	28	0.47	131.00	0
131	Methionine (%)	0227	0.40000	0.00000	0.38524	0.02371	0.00984	28	0.62	131.00	0
131	Methionine (%)	0619	0.40300	0.00200	0.38524	0.02371	0.00984	28	0.75	131.00	0
131	Methionine (%)	0652	0.40500	0.01000	0.38524	0.02371	0.00984	28	0.83	131.00	0
131	Methionine (%)	0910	0.40500	0.01000	0.38524	0.02371	0.00984	28	0.83	131.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
131	Methionine (%)	0610	0.40500	0.01000	0.38524	0.02371	0.00984	28	0.83	131.05	0
131	Methionine (%)	0870	0.40945	0.02570	0.38524	0.02371	0.00984	28	1.02	131.00	0
131	Methionine (%)	0644	0.41150	0.00500	0.38524	0.02371	0.00984	28	1.11	131.00	0
131	Methionine (%)	0675	0.42000	0.02000	0.38524	0.02371	0.00984	28	1.47	131.00	0
131	Methionine (%)	2145	0.42000	0.00000	0.38524	0.02371	0.00984	28	1.47	131.01	0
132	Phenylalanine (%)	0878	0.86000	0.02800	0.94700	0.04832	0.01831	26	-1.80	132.00	0
132	Phenylalanine (%)	0148	0.87500	0.07000	0.94700	0.04832	0.01831	26	-1.49	132.05	0
132	Phenylalanine (%)	0941	0.89000	0.02000	0.94700	0.04832	0.01831	26	-1.18	132.00	0
132	Phenylalanine (%)	0904	0.89000	0.00000	0.94700	0.04832	0.01831	26	-1.18	132.99	0
132	Phenylalanine (%)	0684	0.89450	0.03100	0.94700	0.04832	0.01831	26	-1.09	132.00	0
132	Phenylalanine (%)	0626	0.91000	0.01400	0.94700	0.04832	0.01831	26	-0.77	132.05	0
132	Phenylalanine (%)	0910	0.92500	0.03000	0.94700	0.04832	0.01831	26	-0.46	132.00	0
132	Phenylalanine (%)	2059	0.92550	0.00900	0.94700	0.04832	0.01831	26	-0.45	132.00	0
132	Phenylalanine (%)	0676	0.92750	0.05100	0.94700	0.04832	0.01831	26	-0.40	132.05	0
132	Phenylalanine (%)	0870	0.92860	0.03800	0.94700	0.04832	0.01831	26	-0.38	132.00	0
132	Phenylalanine (%)	0644	0.93750	0.00300	0.94700	0.04832	0.01831	26	-0.20	132.00	0
132	Phenylalanine (%)	0098	0.94300	0.00000	0.94700	0.04832	0.01831	26	-0.08	132.02	0
132	Phenylalanine (%)	0652	0.94500	0.01000	0.94700	0.04832	0.01831	26	-0.04	132.00	0
132	Phenylalanine (%)	0675	0.94500	0.01000	0.94700	0.04832	0.01831	26	-0.04	132.00	0
132	Phenylalanine (%)	0171	0.94950	0.00300	0.94700	0.04832	0.01831	26	0.05	132.00	0
132	Phenylalanine (%)	0571	0.95350	0.00700	0.94700	0.04832	0.01831	26	0.13	132.00	0
132	Phenylalanine (%)	0227	0.96000	0.00000	0.94700	0.04832	0.01831	26	0.27	132.00	0
132	Phenylalanine (%)	0354	0.97350	0.00300	0.94700	0.04832	0.01831	26	0.55	132.00	0
132	Phenylalanine (%)	0918	0.97850	0.00300	0.94700	0.04832	0.01831	26	0.65	132.00	0
132	Phenylalanine (%)	0968	0.98500	0.00800	0.94700	0.04832	0.01831	26	0.79	132.00	0
132	Phenylalanine (%)	0610	0.98500	0.03000	0.94700	0.04832	0.01831	26	0.79	132.05	0
132	Phenylalanine (%)	0859	0.99000	0.01600	0.94700	0.04832	0.01831	26	0.89	132.00	0
132	Phenylalanine (%)	0872	0.99200	0.02200	0.94700	0.04832	0.01831	26	0.93	132.00	0
132	Phenylalanine (%)	0504	1.0050	0.05000	0.94700	0.04832	0.01831	26	1.20	132.00	0
132	Phenylalanine (%)	0955	1.0200	0.02000	0.94700	0.04832	0.01831	26	1.51	132.00	0
132	Phenylalanine (%)	0619	1.0400	0.00000	0.94700	0.04832	0.01831	26	1.92	132.00	0
133	Proline (%)	0910	1.0000	0.02000	1.0693	0.05225	0.02400	25	-1.33	133.00	0
133	Proline (%)	0652	1.0050	0.03000	1.0693	0.05225	0.02400	25	-1.23	133.00	0
133	Proline (%)	0941	1.0200	0.04000	1.0693	0.05225	0.02400	25	-0.94	133.00	0
133	Proline (%)	0872	1.0255	0.01100	1.0693	0.05225	0.02400	25	-0.84	133.00	0
133	Proline (%)	0878	1.0270	0.01200	1.0693	0.05225	0.02400	25	-0.81	133.00	0
133	Proline (%)	0148	1.0300	0.04000	1.0693	0.05225	0.02400	25	-0.75	133.05	0
133	Proline (%)	0684	1.0340	0.01000	1.0693	0.05225	0.02400	25	-0.68	133.00	0
133	Proline (%)	0859	1.0380	0.00800	1.0693	0.05225	0.02400	25	-0.60	133.00	0
133	Proline (%)	0171	1.0405	0.00700	1.0693	0.05225	0.02400	25	-0.55	133.00	0
133	Proline (%)	0870	1.0494	0.01600	1.0693	0.05225	0.02400	25	-0.38	133.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
133	Proline (%)	0626	1.0520	0.02600	1.0693	0.05225	0.02400	25	-0.33	133.05	0
133	Proline (%)	0227	1.0600	0.04000	1.0693	0.05225	0.02400	25	-0.18	133.00	0
133	Proline (%)	0610	1.0600	0.00000	1.0693	0.05225	0.02400	25	-0.18	133.05	0
133	Proline (%)	0904	1.0600	0.06000	1.0693	0.05225	0.02400	25	-0.18	133.99	0
133	Proline (%)	0644	1.0635	0.01100	1.0693	0.05225	0.02400	25	-0.11	133.00	0
133	Proline (%)	0354	1.0665	0.00700	1.0693	0.05225	0.02400	25	-0.05	133.00	0
133	Proline (%)	0571	1.0870	0.00800	1.0693	0.05225	0.02400	25	0.34	133.00	0
133	Proline (%)	2059	1.0915	0.01500	1.0693	0.05225	0.02400	25	0.42	133.00	0
133	Proline (%)	0968	1.0970	0.00400	1.0693	0.05225	0.02400	25	0.53	133.00	0
133	Proline (%)	0675	1.1200	0.02000	1.0693	0.05225	0.02400	25	0.97	133.00	0
133	Proline (%)	0955	1.1300	0.02000	1.0693	0.05225	0.02400	25	1.16	133.00	0
133	Proline (%)	0676	1.1405	0.06100	1.0693	0.05225	0.02400	25	1.36	133.05	0
133	Proline (%)	0918	1.1406	0.01400	1.0693	0.05225	0.02400	25	1.36	133.00	0
133	Proline (%)	0504	1.1750	0.09000	1.0693	0.05225	0.02400	25	2.02	133.00	0
133	Proline (%)	0619	1.4250	0.03000	1.0693	0.05225	0.02400	25	6.81	133.00	0
134	Serine (%)	0504	0.85500	0.01000	0.99556	0.03887	0.01381	26	-3.62	134.00	0
134	Serine (%)	0098	0.92300	0.00400	0.99556	0.03887	0.01381	26	-1.87	134.02	0
134	Serine (%)	0619	0.93900	0.00400	0.99556	0.03887	0.01381	26	-1.46	134.00	0
134	Serine (%)	0941	0.94500	0.05000	0.99556	0.03887	0.01381	26	-1.30	134.00	0
134	Serine (%)	0644	0.95900	0.00400	0.99556	0.03887	0.01381	26	-0.94	134.00	0
134	Serine (%)	0955	0.96750	0.00100	0.99556	0.03887	0.01381	26	-0.72	134.00	0
134	Serine (%)	0626	0.97100	0.02400	0.99556	0.03887	0.01381	26	-0.63	134.05	0
134	Serine (%)	0910	0.97500	0.03000	0.99556	0.03887	0.01381	26	-0.53	134.00	0
134	Serine (%)	0610	0.98500	0.03000	0.99556	0.03887	0.01381	26	-0.27	134.05	0
134	Serine (%)	0878	0.98950	0.00500	0.99556	0.03887	0.01381	26	-0.16	134.00	0
134	Serine (%)	0652	0.99000	0.00000	0.99556	0.03887	0.01381	26	-0.14	134.00	0
134	Serine (%)	0872	0.99000	0.01400	0.99556	0.03887	0.01381	26	-0.14	134.00	0
134	Serine (%)	0354	0.99500	0.01000	0.99556	0.03887	0.01381	26	-0.01	134.00	0
134	Serine (%)	2059	0.99500	0.00800	0.99556	0.03887	0.01381	26	-0.01	134.00	0
134	Serine (%)	0571	1.0095	0.00500	0.99556	0.03887	0.01381	26	0.36	134.00	0
134	Serine (%)	0227	1.0100	0.02000	0.99556	0.03887	0.01381	26	0.37	134.00	0
134	Serine (%)	0859	1.0105	0.00100	0.99556	0.03887	0.01381	26	0.38	134.00	0
134	Serine (%)	0675	1.0150	0.01000	0.99556	0.03887	0.01381	26	0.50	134.00	0
134	Serine (%)	0968	1.0160	0.00600	0.99556	0.03887	0.01381	26	0.53	134.00	0
134	Serine (%)	0918	1.0170	0.00910	0.99556	0.03887	0.01381	26	0.55	134.00	0
134	Serine (%)	0676	1.0275	0.03300	0.99556	0.03887	0.01381	26	0.82	134.05	0
134	Serine (%)	0870	1.0307	0.05500	0.99556	0.03887	0.01381	26	0.90	134.00	0
134	Serine (%)	0684	1.0330	0.00200	0.99556	0.03887	0.01381	26	0.96	134.00	0
134	Serine (%)	0148	1.0400	0.02000	0.99556	0.03887	0.01381	26	1.14	134.05	0
134	Serine (%)	0171	1.0460	0.00400	0.99556	0.03887	0.01381	26	1.30	134.00	0
134	Serine (%)	0904	1.0800	0.00000	0.99556	0.03887	0.01381	26	2.17	134.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
135	Threonine (%)	2145	0.63000	0.02000	0.84981	0.02393	0.01332	28	-9.19	135.01	0
135	Threonine (%)	0160	0.68000	0.00000	0.84981	0.02393	0.01332	28	-7.10	135.99	0
135	Threonine (%)	0626	0.79900	0.01400	0.84981	0.02393	0.01332	28	-2.12	135.05	0
135	Threonine (%)	0941	0.81000	0.04000	0.84981	0.02393	0.01332	28	-1.66	135.00	0
135	Threonine (%)	0148	0.82500	0.03000	0.84981	0.02393	0.01332	28	-1.04	135.05	0
135	Threonine (%)	0619	0.83250	0.00100	0.84981	0.02393	0.01332	28	-0.72	135.00	0
135	Threonine (%)	0504	0.84000	0.02000	0.84981	0.02393	0.01332	28	-0.41	135.00	0
135	Threonine (%)	0904	0.84000	0.00000	0.84981	0.02393	0.01332	28	-0.41	135.99	0
135	Threonine (%)	0872	0.84150	0.00100	0.84981	0.02393	0.01332	28	-0.35	135.00	0
135	Threonine (%)	2059	0.84300	0.00000	0.84981	0.02393	0.01332	28	-0.28	135.00	0
135	Threonine (%)	0644	0.84400	0.00400	0.84981	0.02393	0.01332	28	-0.24	135.00	0
135	Threonine (%)	0859	0.84450	0.00900	0.84981	0.02393	0.01332	28	-0.22	135.00	0
135	Threonine (%)	0878	0.84450	0.00700	0.84981	0.02393	0.01332	28	-0.22	135.00	0
135	Threonine (%)	0684	0.84950	0.03300	0.84981	0.02393	0.01332	28	-0.01	135.00	0
135	Threonine (%)	0955	0.85350	0.02500	0.84981	0.02393	0.01332	28	0.15	135.00	0
135	Threonine (%)	0652	0.85500	0.01000	0.84981	0.02393	0.01332	28	0.22	135.00	0
135	Threonine (%)	0610	0.85500	0.01000	0.84981	0.02393	0.01332	28	0.22	135.05	0
135	Threonine (%)	0676	0.85900	0.04000	0.84981	0.02393	0.01332	28	0.38	135.05	0
135	Threonine (%)	0571	0.86000	0.00600	0.84981	0.02393	0.01332	28	0.43	135.00	0
135	Threonine (%)	0910	0.86000	0.00000	0.84981	0.02393	0.01332	28	0.43	135.00	0
135	Threonine (%)	0354	0.86200	0.00600	0.84981	0.02393	0.01332	28	0.51	135.00	0
135	Threonine (%)	0227	0.86500	0.01000	0.84981	0.02393	0.01332	28	0.63	135.00	0
135	Threonine (%)	0870	0.87110	0.04880	0.84981	0.02393	0.01332	28	0.89	135.00	0
135	Threonine (%)	0968	0.87200	0.00800	0.84981	0.02393	0.01332	28	0.93	135.00	0
135	Threonine (%)	0675	0.87500	0.01000	0.84981	0.02393	0.01332	28	1.05	135.00	0
135	Threonine (%)	0098	0.87550	0.01500	0.84981	0.02393	0.01332	28	1.07	135.02	0
135	Threonine (%)	0171	0.89600	0.00400	0.84981	0.02393	0.01332	28	1.93	135.00	0
135	Threonine (%)	0918	0.90305	0.00110	0.84981	0.02393	0.01332	28	2.22	135.00	0
136	Tryptophan (%)	0955	0.17000	0.00200	0.23357	0.02551	0.00537	16	-2.49	136.03	0
136	Tryptophan (%)	0859	0.19900	0.00200	0.23357	0.02551	0.00537	16	-1.35	136.03	0
136	Tryptophan (%)	0870	0.20370	0.00600	0.23357	0.02551	0.00537	16	-1.17	136.00	0
136	Tryptophan (%)	0619	0.21600	0.01000	0.23357	0.02551	0.00537	16	-0.69	136.00	0
136	Tryptophan (%)	0904	0.22000	0.00000	0.23357	0.02551	0.00537	16	-0.53	136.99	0
136	Tryptophan (%)	0684	0.22800	0.00200	0.23357	0.02551	0.00537	16	-0.22	136.00	0
136	Tryptophan (%)	0878	0.23100	0.00200	0.23357	0.02551	0.00537	16	-0.10	136.01	0
136	Tryptophan (%)	2059	0.23450	0.00500	0.23357	0.02551	0.00537	16	0.04	136.01	0
136	Tryptophan (%)	0918	0.23665	0.01090	0.23357	0.02551	0.00537	16	0.12	136.00	0
136	Tryptophan (%)	0098	0.24000	0.01200	0.23357	0.02551	0.00537	16	0.25	136.02	0
136	Tryptophan (%)	0872	0.24350	0.00100	0.23357	0.02551	0.00537	16	0.39	136.03	0
136	Tryptophan (%)	0571	0.24600	0.00200	0.23357	0.02551	0.00537	16	0.49	136.03	0
136	Tryptophan (%)	0610	0.25000	0.00000	0.23357	0.02551	0.00537	16	0.64	136.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
136	Tryptophan (%)	0644	0.25850	0.00100	0.23357	0.02551	0.00537	16	0.98	136.01	0
136	Tryptophan (%)	0504	0.26500	0.01000	0.23357	0.02551	0.00537	16	1.23	136.99	0
136	Tryptophan (%)	0227	0.27000	0.02000	0.23357	0.02551	0.00537	16	1.43	136.00	0
137	Tyrosine (%)	0870	0.38625	0.01290	0.63446	0.07132	0.02476	20	-3.48	137.00	0
137	Tyrosine (%)	0910	0.47000	0.08000	0.63446	0.07132	0.02476	20	-2.31	137.00	0
137	Tyrosine (%)	0684	0.56450	0.02900	0.63446	0.07132	0.02476	20	-0.98	137.00	0
137	Tyrosine (%)	0626	0.57200	0.04000	0.63446	0.07132	0.02476	20	-0.88	137.05	0
137	Tyrosine (%)	0148	0.58500	0.05000	0.63446	0.07132	0.02476	20	-0.69	137.05	0
137	Tyrosine (%)	0941	0.59000	0.08000	0.63446	0.07132	0.02476	20	-0.62	137.00	0
137	Tyrosine (%)	0675	0.60000	0.00000	0.63446	0.07132	0.02476	20	-0.48	137.00	0
137	Tyrosine (%)	0354	0.61800	0.00400	0.63446	0.07132	0.02476	20	-0.23	137.00	0
137	Tyrosine (%)	0227	0.62500	0.01000	0.63446	0.07132	0.02476	20	-0.13	137.00	0
137	Tyrosine (%)	0918	0.62575	0.01230	0.63446	0.07132	0.02476	20	-0.12	137.00	0
137	Tyrosine (%)	0171	0.63900	0.00600	0.63446	0.07132	0.02476	20	0.06	137.00	0
137	Tyrosine (%)	0872	0.64050	0.01100	0.63446	0.07132	0.02476	20	0.08	137.00	0
137	Tyrosine (%)	0644	0.64650	0.01100	0.63446	0.07132	0.02476	20	0.17	137.00	0
137	Tyrosine (%)	0968	0.66550	0.00700	0.63446	0.07132	0.02476	20	0.44	137.00	0
137	Tyrosine (%)	0676	0.68650	0.02900	0.63446	0.07132	0.02476	20	0.73	137.05	0
137	Tyrosine (%)	2059	0.70350	0.00700	0.63446	0.07132	0.02476	20	0.97	137.00	0
137	Tyrosine (%)	0504	0.70500	0.05000	0.63446	0.07132	0.02476	20	0.99	137.00	0
137	Tyrosine (%)	0904	0.71000	0.04000	0.63446	0.07132	0.02476	20	1.06	137.99	0
137	Tyrosine (%)	0098	0.71600	0.01600	0.63446	0.07132	0.02476	20	1.14	137.02	0
137	Tyrosine (%)	0610	0.75000	0.00000	0.63446	0.07132	0.02476	20	1.62	137.05	0
138	Valine (%)	0148	0.86500	0.01000	0.97449	0.04909	0.02367	26	-2.23	138.05	0
138	Valine (%)	0941	0.91000	0.04000	0.97449	0.04909	0.02367	26	-1.31	138.00	0
138	Valine (%)	0171	0.91250	0.00500	0.97449	0.04909	0.02367	26	-1.26	138.00	0
138	Valine (%)	0626	0.91600	0.04400	0.97449	0.04909	0.02367	26	-1.19	138.05	0
138	Valine (%)	0859	0.91850	0.00900	0.97449	0.04909	0.02367	26	-1.14	138.00	0
138	Valine (%)	0918	0.93120	0.03080	0.97449	0.04909	0.02367	26	-0.88	138.00	0
138	Valine (%)	0684	0.93750	0.07900	0.97449	0.04909	0.02367	26	-0.75	138.00	0
138	Valine (%)	0652	0.94000	0.02000	0.97449	0.04909	0.02367	26	-0.70	138.00	0
138	Valine (%)	0870	0.94900	0.01360	0.97449	0.04909	0.02367	26	-0.52	138.00	0
138	Valine (%)	0872	0.95300	0.01800	0.97449	0.04909	0.02367	26	-0.44	138.00	0
138	Valine (%)	0676	0.95650	0.06300	0.97449	0.04909	0.02367	26	-0.37	138.05	0
138	Valine (%)	0354	0.98100	0.00600	0.97449	0.04909	0.02367	26	0.13	138.00	0
138	Valine (%)	2059	0.98400	0.00200	0.97449	0.04909	0.02367	26	0.19	138.00	0
138	Valine (%)	0675	0.98500	0.01000	0.97449	0.04909	0.02367	26	0.21	138.00	0
138	Valine (%)	0904	0.99500	0.01000	0.97449	0.04909	0.02367	26	0.42	138.99	0
138	Valine (%)	0610	1.0000	0.02000	0.97449	0.04909	0.02367	26	0.52	138.05	0
138	Valine (%)	0878	1.0015	0.02300	0.97449	0.04909	0.02367	26	0.55	138.00	0
138	Valine (%)	0968	1.0040	0.02600	0.97449	0.04909	0.02367	26	0.60	138.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
138	Valine (%)	0571	1.0080	0.01600	0.97449	0.04909	0.02367	26	0.68	138.00	0
138	Valine (%)	0644	1.0100	0.00200	0.97449	0.04909	0.02367	26	0.72	138.00	0
138	Valine (%)	0910	1.0100	0.06000	0.97449	0.04909	0.02367	26	0.72	138.00	0
138	Valine (%)	0619	1.0120	0.03600	0.97449	0.04909	0.02367	26	0.76	138.00	0
138	Valine (%)	0227	1.0150	0.03000	0.97449	0.04909	0.02367	26	0.83	138.00	0
138	Valine (%)	0504	1.0250	0.01000	0.97449	0.04909	0.02367	26	1.03	138.00	0
138	Valine (%)	0098	1.0330	0.00200	0.97449	0.04909	0.02367	26	1.19	138.02	0
138	Valine (%)	0955	1.1950	0.03000	0.97449	0.04909	0.02367	26	4.49	138.00	0
139	Taurine (%)	0148	0.02000	0.00000	0.02863	0.00594	0.00175	4	-1.45	139.05	0
139	Taurine (%)	0227	0.03000	0.00000	0.02863	0.00594	0.00175	4	0.23	139.00	0
139	Taurine (%)	0171	0.03100	0.00200	0.02863	0.00594	0.00175	4	0.40	139.00	0
139	Taurine (%)	0098	0.03350	0.00500	0.02863	0.00594	0.00175	4	0.82	139.02	0
150	Phytase (Units / kg)	2068	1,285.0	10.000	1,694.9	305.33	107.00	5	-1.34	150.99	0
150	Phytase (Units / kg)	0610	1,527.5	185.00	1,694.9	305.33	107.00	5	-0.55	150.00	0
150	Phytase (Units / kg)	0904	1,700.0	140.00	1,694.9	305.33	107.00	5	0.02	150.99	0
150	Phytase (Units / kg)	0171	1,900.0	200.00	1,694.9	305.33	107.00	5	0.67	150.00	0
150	Phytase (Units / kg)	0619	2,062.0	0.00000	1,694.9	305.33	107.00	5	1.20	150.00	0
160	Fructose (%)	0619	0.34300	0.00600	0.58960	0.28422	0.13520	5	-0.94	160.99	0
160	Fructose (%)	0227	0.38500	0.03000	0.58960	0.28422	0.13520	5	-0.64	160.99	0
160	Fructose (%)	0904	0.52000	0.04000	0.58960	0.28422	0.13520	5	0.33	160.99	0
160	Fructose (%)	0861	0.65000	0.10000	0.58960	0.28422	0.13520	5	1.26	160.99	0
160	Fructose (%)	0910	1.0500	0.50000	0.58960	0.28422	0.13520	5	4.13	160.99	0
161	Galactose (%)	0904	0.16500	0.01000				1		161.99	0
162	Glucose (%)	0227	0.51000	0.00000	0.85830	0.48288	0.21940	5	-1.10	162.99	0
162	Glucose (%)	0619	0.59150	0.01700	0.85830	0.48288	0.21940	5	-0.45	162.99	0
162	Glucose (%)	0904	0.69000	0.08000	0.85830	0.48288	0.21940	5	0.34	162.99	0
162	Glucose (%)	0861	0.80000	0.20000	0.85830	0.48288	0.21940	5	1.21	162.99	0
162	Glucose (%)	0910	1.7000	0.80000	0.85830	0.48288	0.21940	5	8.40	162.99	0
163	Lactose (%)	0861	15.100	0.40000	16.711	1.4572	0.23240	10	-1.11	163.99	0
163	Lactose (%)	0910	15.500	0.20000	16.711	1.4572	0.23240	10	-0.83	163.99	0
163	Lactose (%)	0148	15.575	0.03000	16.711	1.4572	0.23240	10	-0.78	163.99	0
163	Lactose (%)	0619	16.100	0.00000	16.711	1.4572	0.23240	10	-0.42	163.99	0
163	Lactose (%)	0027	16.340	0.27400	16.711	1.4572	0.23240	10	-0.25	163.99	0
163	Lactose (%)	0629	16.375	0.25000	16.711	1.4572	0.23240	10	-0.23	163.99	0
163	Lactose (%)	0227	17.045	0.93000	16.711	1.4572	0.23240	10	0.23	163.99	0
163	Lactose (%)	0904	17.780	0.04000	16.711	1.4572	0.23240	10	0.73	163.99	0
163	Lactose (%)	0689	18.400	0.00000	16.711	1.4572	0.23240	10	1.16	163.99	0
163	Lactose (%)	2129	19.200	0.20000	16.711	1.4572	0.23240	10	1.71	163.99	0
164	Maltose (%)	0227	0.57500	0.27000				3		164.99	0
164	Maltose (%)	0904	2.5850	0.21000				3		164.99	0
164	Maltose (%)	0910	3.7500	0.30000				3		164.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
164	Maltose (%)	0861	0.00000	0.00000				3		164.99	4
165	Sucrose (%)	0910	3.6500	0.50000	4.1025	0.39972	0.24000	4	-1.13	165.99	0
165	Sucrose (%)	0227	3.8850	0.21000	4.1025	0.39972	0.24000	4	-0.54	165.99	0
165	Sucrose (%)	0861	4.4000	0.20000	4.1025	0.39972	0.24000	4	0.74	165.99	0
165	Sucrose (%)	0619	4.4750	0.05000	4.1025	0.39972	0.24000	4	0.93	165.99	0
166	Raffinose (%)	0227	0.17500	0.01000				2		166.99	0
166	Raffinose (%)	0619	0.24500	0.01000				2		166.99	0
167	Stachyose (%)	0227	0.79500	0.01000				2		167.99	0
167	Stachyose (%)	0619	1.2250	0.05000				2		167.99	0
350	Carbadox (mg/kg (ppm))	2053	0.22500	0.17000	253.19	504.67	4.9175	4	-0.95	350.02	0
350	Carbadox (mg/kg (ppm))	2149	0.80000	0.00000	253.19	504.67	4.9175	4	-0.09	350.01	0
350	Carbadox (mg/kg (ppm))	0047	1.5500	0.70000	253.19	504.67	4.9175	4	1.04	350.01	0
350	Carbadox (mg/kg (ppm))	2135	1,010.2	18.800	253.19	504.67	4.9175	4	1519.12	350.03	0
351	Chlortetracycline (mg/kg (ppm))	2135	839.30	136.60				1		351.05	0
357	Ethoxyquin (mg/kg (ppm))	0942	50.500	1.0000				1		357.99	0
373	Oxytetracycline (mg/kg (ppm))	0047	262.00	22.000	346.57	56.856	18.991	17	-1.49	373.04	0
373	Oxytetracycline (mg/kg (ppm))	2145	292.70	24.000	346.57	56.856	18.991	17	-0.95	373.05	0
373	Oxytetracycline (mg/kg (ppm))	0001	297.50	5.0000	346.57	56.856	18.991	17	-0.86	373.04	0
373	Oxytetracycline (mg/kg (ppm))	0036	306.90	10.000	346.57	56.856	18.991	17	-0.70	373.03	0
373	Oxytetracycline (mg/kg (ppm))	0033	309.00	4.0000	346.57	56.856	18.991	17	-0.66	373.04	0
373	Oxytetracycline (mg/kg (ppm))	0027	313.91	11.313	346.57	56.856	18.991	17	-0.57	373.00	0
373	Oxytetracycline (mg/kg (ppm))	0218	321.33	8.8770	346.57	56.856	18.991	17	-0.44	373.03	0
373	Oxytetracycline (mg/kg (ppm))	0013	327.00	2.0000	346.57	56.856	18.991	17	-0.34	373.04	0
373	Oxytetracycline (mg/kg (ppm))	0964	332.75	31.300	346.57	56.856	18.991	17	-0.24	373.00	0
373	Oxytetracycline (mg/kg (ppm))	2053	350.00	8.0000	346.57	56.856	18.991	17	0.06	373.05	0
373	Oxytetracycline (mg/kg (ppm))	0910	362.00	20.000	346.57	56.856	18.991	17	0.27	373.03	0
373	Oxytetracycline (mg/kg (ppm))	0917	369.00	56.000	346.57	56.856	18.991	17	0.39	373.03	0
373	Oxytetracycline (mg/kg (ppm))	0689	387.30	0.00000	346.57	56.856	18.991	17	0.72	373.03	0
373	Oxytetracycline (mg/kg (ppm))	0511	388.50	61.000	346.57	56.856	18.991	17	0.74	373.00	0
373	Oxytetracycline (mg/kg (ppm))	0028	408.05	22.100	346.57	56.856	18.991	17	1.08	373.06	0
373	Oxytetracycline (mg/kg (ppm))	0035	441.43	20.460	346.57	56.856	18.991	17	1.67	373.00	0
373	Oxytetracycline (mg/kg (ppm))	0904	443.57	16.800	346.57	56.856	18.991	17	1.71	373.03	0
373	Oxytetracycline (mg/kg (ppm))	2127	266.00	100.00	346.57	56.856	18.991	17	-1.42	373.06	1
377	Pyrantel Tartrate (mg/kg (ppm))	2053	2.2500	0.10000				1		377.02	0
386	Tiamulin (mg/kg (ppm))	0875	0.32000	0.00000				2		386.02	0
386	Tiamulin (mg/kg (ppm))	2053	0.55000	0.04000				2		386.01	0
393	Ractopamine Hydrochloride (mg/kg (ppn	2149	0.15665	0.01930				3		393.99	0
393	Ractopamine Hydrochloride (mg/kg (ppn	2145	135.00	6.0000				3		393.02	0
393	Ractopamine Hydrochloride (mg/kg (ppn	0875	271.50	23.000				3		393.02	0
400	Water activity (Units)	2073	0.56850	0.00700	0.59506	0.01817	0.00297	9	-1.46	400.99	0
400	Water activity (Units)	0843	0.57050	0.00900	0.59506	0.01817	0.00297	9	-1.35	400.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
400	Water activity (Units)	0843	0.59205	0.00010	0.59506	0.01817	0.00297	9	-0.17	400.01	0
400	Water activity (Units)	0407	0.59495	0.00110	0.59506	0.01817	0.00297	9	-0.01	400.01	0
400	Water activity (Units)	0942	0.59550	0.00300	0.59506	0.01817	0.00297	9	0.02	400.01	0
400	Water activity (Units)	0010	0.60650	0.00100	0.59506	0.01817	0.00297	9	0.63	400.01	0
400	Water activity (Units)	0589	0.60800	0.00200	0.59506	0.01817	0.00297	9	0.71	400.01	0
400	Water activity (Units)	2109	0.60830	0.00120	0.59506	0.01817	0.00297	9	0.73	400.01	0
400	Water activity (Units)	0552	0.61125	0.00230	0.59506	0.01817	0.00297	9	0.89	400.01	0
412		0353	18.765	0.03000				1		412.01	0
516	Arsenic, total (mg / kg (ppm))	0045	0.13950	0.00700	0.32716	0.06748	0.02040	9	-2.78	516.00	0
516	Arsenic, total (mg / kg (ppm))	2146	0.26045	0.03350	0.32716	0.06748	0.02040	9	-0.99	516.52	0
516	Arsenic, total (mg / kg (ppm))	0171	0.29500	0.01000	0.32716	0.06748	0.02040	9	-0.48	516.00	0
516	Arsenic, total (mg / kg (ppm))	0186	0.32750	0.00300	0.32716	0.06748	0.02040	9	0.01	516.52	0
516	Arsenic, total (mg / kg (ppm))	0227	0.33850	0.01100	0.32716	0.06748	0.02040	9	0.17	516.53	0
516	Arsenic, total (mg / kg (ppm))	0553	0.35300	0.01800	0.32716	0.06748	0.02040	9	0.38	516.53	0
516	Arsenic, total (mg / kg (ppm))	0918	0.35565	0.01610	0.32716	0.06748	0.02040	9	0.42	516.53	0
516	Arsenic, total (mg / kg (ppm))	0910	0.36000	0.02000	0.32716	0.06748	0.02040	9	0.49	516.52	0
516	Arsenic, total (mg / kg (ppm))	0098	0.46650	0.06500	0.32716	0.06748	0.02040	9	2.06	516.53	0
518	Cadmium (mg / kg (ppm))	0186	0.10450	0.00700	0.16316	0.01560	0.00641	9	-3.76	518.52	0
518	Cadmium (mg / kg (ppm))	0918	0.15070	0.00060	0.16316	0.01560	0.00641	9	-0.80	518.53	0
518	Cadmium (mg / kg (ppm))	0227	0.15550	0.01100	0.16316	0.01560	0.00641	9	-0.49	518.53	0
518	Cadmium (mg / kg (ppm))	2055	0.15900	0.00400	0.16316	0.01560	0.00641	9	-0.27	518.53	0
518	Cadmium (mg / kg (ppm))	0098	0.16300	0.01200	0.16316	0.01560	0.00641	9	-0.01	518.53	0
518	Cadmium (mg / kg (ppm))	0553	0.16850	0.00300	0.16316	0.01560	0.00641	9	0.34	518.53	0
518	Cadmium (mg / kg (ppm))	2146	0.17045	0.01010	0.16316	0.01560	0.00641	9	0.47	518.52	0
518	Cadmium (mg / kg (ppm))	0968	0.17500	0.01000	0.16316	0.01560	0.00641	9	0.76	518.43	0
518	Cadmium (mg / kg (ppm))	0910	0.25000	0.00000	0.16316	0.01560	0.00641	9	5.57	518.52	0
518	Cadmium (mg / kg (ppm))	0171	0.11000	0.04000	0.16316	0.01560	0.00641	9	-3.41	518.41	1
520	Chromium (mg / kg (ppm))	0186	2.0900	0.06000	5.1746	3.3652	0.55845	8	-0.92	520.52	0
520	Chromium (mg / kg (ppm))	2055	3.0135	0.51500	5.1746	3.3652	0.55845	8	-0.64	520.53	0
520	Chromium (mg / kg (ppm))	0918	3.1327	0.10880	5.1746	3.3652	0.55845	8	-0.61	520.53	0
520	Chromium (mg / kg (ppm))	0553	3.9950	1.2900	5.1746	3.3652	0.55845	8	-0.35	520.53	0
520	Chromium (mg / kg (ppm))	0171	4.6000	0.20000	5.1746	3.3652	0.55845	8	-0.17	520.41	0
520	Chromium (mg / kg (ppm))	0968	5.1000	0.46000	5.1746	3.3652	0.55845	8	-0.02	520.43	0
520	Chromium (mg / kg (ppm))	0510	9.6700	0.18000	5.1746	3.3652	0.55845	8	1.34	520.43	0
520	Chromium (mg / kg (ppm))	0693	9.7956	1.6538	5.1746	3.3652	0.55845	8	1.37	520.42	0
526	Lead (mg / kg (ppm))	2146	0.36050	0.07700	0.60030	0.15377	0.02274	8	-1.56	526.52	0
526	Lead (mg / kg (ppm))	0171	0.40000	0.00000	0.60030	0.15377	0.02274	8	-1.30	526.41	0
526	Lead (mg / kg (ppm))	0227	0.63000	0.01800	0.60030	0.15377	0.02274	8	0.19	526.53	0
526	Lead (mg / kg (ppm))	0918	0.65775	0.01490	0.60030	0.15377	0.02274	8	0.37	526.53	0
526	Lead (mg / kg (ppm))	0553	0.67000	0.02600	0.60030	0.15377	0.02274	8	0.45	526.53	0
526	Lead (mg / kg (ppm))	0098	0.67050	0.00500	0.60030	0.15377	0.02274	8	0.46	526.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
526	Lead (mg / kg (ppm))	0186	0.68450	0.02100	0.60030	0.15377	0.02274	8	0.55	526.52	0
526	Lead (mg / kg (ppm))	0910	0.72000	0.02000	0.60030	0.15377	0.02274	8	0.78	526.52	0
539	Nickel (mg / kg (ppm))	0186	2.5650	0.09000	2.7996	0.39520	0.13745	4	-0.87	539.52	0
539	Nickel (mg / kg (ppm))	0918	2.5933	0.11980	2.7996	0.39520	0.13745	4	-0.22	539.53	0
539	Nickel (mg / kg (ppm))	0171	2.6500	0.30000	2.7996	0.39520	0.13745	4	1.09	539.41	0
539	Nickel (mg / kg (ppm))	2053	3.3900	0.04000	2.7996	0.39520	0.13745	4	18.19	539.42	0

Notes: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2 and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = Rejected for duplicates too far apart, 2 = Rejected as outlier, 8 = Analyst data exempt and 4 = zeros submitted as values. Robust statistics not used if < 6 labs reporting, in this case the Z Scores may be included for information only (Grey, No Action!). Flag 3 indicates not used in statistics.