



Fish Food, Carnivore

Animal Feed Scheme

Labs Reporting: 195

Test Material Code # 201732

Analyte Proficiency Testing Report

Issue Date : 01/31/2018

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
000	Urea (%)	0278	0.50000	0.00000				1		000.02	0
001	Loss on Drying (%)	0618	3.3000	0.18000	4.8925	0.21016	0.08794	76	-7.58	001.99	0
001	Loss on Drying (%)	0618	3.7250	0.01000	4.8925	0.21016	0.08794	76	-5.56	001.07	0
001	Loss on Drying (%)	0017	4.0000	0.20000	4.8925	0.21016	0.08794	76	-4.25	001.00	0
001	Loss on Drying (%)	0019	4.1000	0.20000	4.8925	0.21016	0.08794	76	-3.77	001.07	0
001	Loss on Drying (%)	0720	4.1250	0.01000	4.8925	0.21016	0.08794	76	-3.65	001.99	0
001	Loss on Drying (%)	0536	4.2700	0.06000	4.8925	0.21016	0.08794	76	-2.96	001.99	0
001	Loss on Drying (%)	2155	4.3550	0.27000	4.8925	0.21016	0.08794	76	-2.56	001.99	0
001	Loss on Drying (%)	0876	4.3850	0.05000	4.8925	0.21016	0.08794	76	-2.41	001.07	0
001	Loss on Drying (%)	0560	4.3950	0.13000	4.8925	0.21016	0.08794	76	-2.37	001.99	0
001	Loss on Drying (%)	0038	4.4850	0.23000	4.8925	0.21016	0.08794	76	-1.94	001.07	0
001	Loss on Drying (%)	0034	4.6000	0.26000	4.8925	0.21016	0.08794	76	-1.39	001.99	0
001	Loss on Drying (%)	0510	4.6500	0.10000	4.8925	0.21016	0.08794	76	-1.15	001.99	0
001	Loss on Drying (%)	0016	4.6750	0.03000	4.8925	0.21016	0.08794	76	-1.03	001.00	0
001	Loss on Drying (%)	0918	4.7050	0.23000	4.8925	0.21016	0.08794	76	-0.89	001.99	0
001	Loss on Drying (%)	0676	4.7100	0.02000	4.8925	0.21016	0.08794	76	-0.87	001.99	0
001	Loss on Drying (%)	0186	4.7150	0.03000	4.8925	0.21016	0.08794	76	-0.84	001.03	0
001	Loss on Drying (%)	0610	4.7500	0.00000	4.8925	0.21016	0.08794	76	-0.68	001.05	0
001	Loss on Drying (%)	0142	4.7500	0.06000	4.8925	0.21016	0.08794	76	-0.68	001.07	0
001	Loss on Drying (%)	0003	4.7700	0.06000	4.8925	0.21016	0.08794	76	-0.58	001.07	0
001	Loss on Drying (%)	0171	4.7700	0.04000	4.8925	0.21016	0.08794	76	-0.58	001.07	0
001	Loss on Drying (%)	0297	4.7700	0.16000	4.8925	0.21016	0.08794	76	-0.58	001.07	0
001	Loss on Drying (%)	0683	4.7850	0.03000	4.8925	0.21016	0.08794	76	-0.51	001.07	0
001	Loss on Drying (%)	0226	4.8000	0.00000	4.8925	0.21016	0.08794	76	-0.44	001.07	0
001	Loss on Drying (%)	0413	4.8000	0.20000	4.8925	0.21016	0.08794	76	-0.44	001.07	0
001	Loss on Drying (%)	0309	4.8050	0.23000	4.8925	0.21016	0.08794	76	-0.42	001.00	0
001	Loss on Drying (%)	0015	4.8100	0.20000	4.8925	0.21016	0.08794	76	-0.39	001.07	0
001	Loss on Drying (%)	0353	4.8100	0.12000	4.8925	0.21016	0.08794	76	-0.39	001.07	0
001	Loss on Drying (%)	2149	4.8140	0.28000	4.8925	0.21016	0.08794	76	-0.37	001.99	0
001	Loss on Drying (%)	0083	4.8250	0.05000	4.8925	0.21016	0.08794	76	-0.32	001.07	0
001	Loss on Drying (%)	0619	4.8500	0.04000	4.8925	0.21016	0.08794	76	-0.20	001.03	0
001	Loss on Drying (%)	0512	4.8640	0.02400	4.8925	0.21016	0.08794	76	-0.14	001.07	0
001	Loss on Drying (%)	0590	4.8650	0.05000	4.8925	0.21016	0.08794	76	-0.13	001.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
001	Loss on Drying (%)	0849	4.8900	0.02000	4.8925	0.21016	0.08794	76	-0.01	001.07	0
001	Loss on Drying (%)	0139	4.8950	0.01000	4.8925	0.21016	0.08794	76	0.01	001.07	0
001	Loss on Drying (%)	0843	4.8950	0.31000	4.8925	0.21016	0.08794	76	0.01	001.07	0
001	Loss on Drying (%)	0852	4.8950	0.03000	4.8925	0.21016	0.08794	76	0.01	001.99	0
001	Loss on Drying (%)	0689	4.9000	0.00000	4.8925	0.21016	0.08794	76	0.04	001.07	0
001	Loss on Drying (%)	0504	4.9450	0.15000	4.8925	0.21016	0.08794	76	0.25	001.00	0
001	Loss on Drying (%)	2009	4.9491	0.01030	4.8925	0.21016	0.08794	76	0.27	001.07	0
001	Loss on Drying (%)	0037	4.9500	0.10000	4.8925	0.21016	0.08794	76	0.27	001.99	0
001	Loss on Drying (%)	0910	4.9500	0.10000	4.8925	0.21016	0.08794	76	0.27	001.99	0
001	Loss on Drying (%)	0895	4.9550	0.01000	4.8925	0.21016	0.08794	76	0.30	001.03	0
001	Loss on Drying (%)	0581	4.9550	0.07000	4.8925	0.21016	0.08794	76	0.30	001.07	0
001	Loss on Drying (%)	0948	4.9550	0.03000	4.8925	0.21016	0.08794	76	0.30	001.99	0
001	Loss on Drying (%)	0571	4.9600	0.00000	4.8925	0.21016	0.08794	76	0.32	001.07	0
001	Loss on Drying (%)	0693	4.9630	0.02000	4.8925	0.21016	0.08794	76	0.34	001.07	0
001	Loss on Drying (%)	0505	4.9700	0.10000	4.8925	0.21016	0.08794	76	0.37	001.99	0
001	Loss on Drying (%)	2059	4.9850	0.09000	4.8925	0.21016	0.08794	76	0.44	001.07	0
001	Loss on Drying (%)	0027	4.9900	0.00400	4.8925	0.21016	0.08794	76	0.46	001.00	0
001	Loss on Drying (%)	0066	4.9900	0.08000	4.8925	0.21016	0.08794	76	0.46	001.07	0
001	Loss on Drying (%)	0912	5.0000	0.24000	4.8925	0.21016	0.08794	76	0.51	001.99	0
001	Loss on Drying (%)	0035	5.0050	0.11000	4.8925	0.21016	0.08794	76	0.54	001.07	0
001	Loss on Drying (%)	0638	5.0100	0.02000	4.8925	0.21016	0.08794	76	0.56	001.07	0
001	Loss on Drying (%)	2146	5.0150	0.05000	4.8925	0.21016	0.08794	76	0.58	001.07	0
001	Loss on Drying (%)	2053	5.0200	0.00000	4.8925	0.21016	0.08794	76	0.61	001.07	0
001	Loss on Drying (%)	0903	5.0350	0.01000	4.8925	0.21016	0.08794	76	0.68	001.03	0
001	Loss on Drying (%)	0644	5.0350	0.03000	4.8925	0.21016	0.08794	76	0.68	001.05	0
001	Loss on Drying (%)	0098	5.0350	0.11000	4.8925	0.21016	0.08794	76	0.68	001.07	0
001	Loss on Drying (%)	0940	5.0350	0.01000	4.8925	0.21016	0.08794	76	0.68	001.99	0
001	Loss on Drying (%)	0878	5.0400	0.00000	4.8925	0.21016	0.08794	76	0.70	001.03	0
001	Loss on Drying (%)	0169	5.0450	0.01000	4.8925	0.21016	0.08794	76	0.73	001.00	0
001	Loss on Drying (%)	0366	5.0500	0.50000	4.8925	0.21016	0.08794	76	0.75	001.07	0
001	Loss on Drying (%)	2012	5.0550	0.01000	4.8925	0.21016	0.08794	76	0.77	001.07	0
001	Loss on Drying (%)	2062	5.0585	0.01230	4.8925	0.21016	0.08794	76	0.79	001.03	0
001	Loss on Drying (%)	0675	5.0650	0.01000	4.8925	0.21016	0.08794	76	0.82	001.07	0
001	Loss on Drying (%)	0893	5.0700	0.00000	4.8925	0.21016	0.08794	76	0.84	001.03	0
001	Loss on Drying (%)	0049	5.0700	0.20000	4.8925	0.21016	0.08794	76	0.84	001.07	0
001	Loss on Drying (%)	0872	5.1050	0.01000	4.8925	0.21016	0.08794	76	1.01	001.07	0
001	Loss on Drying (%)	0629	5.1150	0.05000	4.8925	0.21016	0.08794	76	1.06	001.99	0
001	Loss on Drying (%)	0407	5.1400	0.06310	4.8925	0.21016	0.08794	76	1.18	001.07	0
001	Loss on Drying (%)	0089	5.1400	0.00000	4.8925	0.21016	0.08794	76	1.18	001.07	0
001	Loss on Drying (%)	0592	5.1950	0.13000	4.8925	0.21016	0.08794	76	1.44	001.07	0
001	Loss on Drying (%)	0357	5.3509	0.07990	4.8925	0.21016	0.08794	76	2.18	001.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
001	Loss on Drying (%)	0278	5.3750	0.13000	4.8925	0.21016	0.08794	76	2.30	001.07	0
001	Loss on Drying (%)	0529	5.3800	0.04000	4.8925	0.21016	0.08794	76	2.32	001.07	0
001	Loss on Drying (%)	0045	6.0250	0.17000	4.8925	0.21016	0.08794	76	5.39	001.07	0
001	Loss on Drying (%)	0630	5.6550	2.1100	4.8925	0.21016	0.08794	76	3.63	001.99	1
002	Protein (%)	0183	43.140	0.36000	51.874	0.61735	0.28276	190	-14.15	002.01	0
002	Protein (%)	0405	47.690	1.9800	51.874	0.61735	0.28276	190	-6.78	002.04	0
002	Protein (%)	0874	48.860	0.16000	51.874	0.61735	0.28276	190	-4.88	002.04	0
002	Protein (%)	0722	49.095	0.37150	51.874	0.61735	0.28276	190	-4.50	002.05	0
002	Protein (%)	0536	49.420	0.00000	51.874	0.61735	0.28276	190	-3.97	002.05	0
002	Protein (%)	0859	49.544	0.48000	51.874	0.61735	0.28276	190	-3.77	002.06	0
002	Protein (%)	0852	49.750	0.30000	51.874	0.61735	0.28276	190	-3.44	002.11	0
002	Protein (%)	0622	50.255	0.06620	51.874	0.61735	0.28276	190	-2.62	002.05	0
002	Protein (%)	0066	50.510	0.14000	51.874	0.61735	0.28276	190	-2.21	002.02	0
002	Protein (%)	0885	50.677	0.25420	51.874	0.61735	0.28276	190	-1.94	002.05	0
002	Protein (%)	2006	50.680	0.12000	51.874	0.61735	0.28276	190	-1.93	002.05	0
002	Protein (%)	0164	50.715	0.33000	51.874	0.61735	0.28276	190	-1.88	002.01	0
002	Protein (%)	2073	50.765	0.17000	51.874	0.61735	0.28276	190	-1.80	002.05	0
002	Protein (%)	2149	50.800	0.24900	51.874	0.61735	0.28276	190	-1.74	002.02	0
002	Protein (%)	2022	50.835	0.01000	51.874	0.61735	0.28276	190	-1.68	002.05	0
002	Protein (%)	0674	50.870	0.00000	51.874	0.61735	0.28276	190	-1.63	002.05	0
002	Protein (%)	0354	50.905	0.03000	51.874	0.61735	0.28276	190	-1.57	002.05	0
002	Protein (%)	0651	50.913	0.10400	51.874	0.61735	0.28276	190	-1.56	002.05	0
002	Protein (%)	0169	50.925	0.87000	51.874	0.61735	0.28276	190	-1.54	002.00	0
002	Protein (%)	2081	50.940	0.42000	51.874	0.61735	0.28276	190	-1.51	002.06	0
002	Protein (%)	0918	50.965	0.75000	51.874	0.61735	0.28276	190	-1.47	002.06	0
002	Protein (%)	2006	50.975	0.21000	51.874	0.61735	0.28276	190	-1.46	002.06	0
002	Protein (%)	0910	51.000	0.80000	51.874	0.61735	0.28276	190	-1.42	002.01	0
002	Protein (%)	0098	51.030	0.18000	51.874	0.61735	0.28276	190	-1.37	002.08	0
002	Protein (%)	2009	51.087	0.00780	51.874	0.61735	0.28276	190	-1.28	002.05	0
002	Protein (%)	0893	51.090	0.00000	51.874	0.61735	0.28276	190	-1.27	002.05	0
002	Protein (%)	0552	51.095	0.03000	51.874	0.61735	0.28276	190	-1.26	002.05	0
002	Protein (%)	0619	51.150	0.10000	51.874	0.61735	0.28276	190	-1.17	002.05	0
002	Protein (%)	0968	51.157	0.06900	51.874	0.61735	0.28276	190	-1.16	002.01	0
002	Protein (%)	0689	51.200	0.20000	51.874	0.61735	0.28276	190	-1.09	002.05	0
002	Protein (%)	0653	51.205	0.01000	51.874	0.61735	0.28276	190	-1.08	002.01	0
002	Protein (%)	2068	51.220	0.04000	51.874	0.61735	0.28276	190	-1.06	002.05	0
002	Protein (%)	0720	51.230	0.22000	51.874	0.61735	0.28276	190	-1.04	002.11	0
002	Protein (%)	2144	51.270	0.08000	51.874	0.61735	0.28276	190	-0.98	002.99	0
002	Protein (%)	0728	51.297	0.23300	51.874	0.61735	0.28276	190	-0.94	002.04	0
002	Protein (%)	0598	51.340	0.74000	51.874	0.61735	0.28276	190	-0.86	002.06	0
002	Protein (%)	0720	51.350	0.16000	51.874	0.61735	0.28276	190	-0.85	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein (%)	0108	51.355	1.4500	51.874	0.61735	0.28276	190	-0.84	002.06	0
002	Protein (%)	2146	51.370	1.0800	51.874	0.61735	0.28276	190	-0.82	002.05	0
002	Protein (%)	0546	51.370	0.21800	51.874	0.61735	0.28276	190	-0.82	002.06	0
002	Protein (%)	2145	51.375	0.13000	51.874	0.61735	0.28276	190	-0.81	002.99	0
002	Protein (%)	0675	51.385	0.11000	51.874	0.61735	0.28276	190	-0.79	002.05	0
002	Protein (%)	0265	51.385	0.41000	51.874	0.61735	0.28276	190	-0.79	002.06	0
002	Protein (%)	0941	51.390	0.46000	51.874	0.61735	0.28276	190	-0.78	002.05	0
002	Protein (%)	0297	51.455	1.2300	51.874	0.61735	0.28276	190	-0.68	002.06	0
002	Protein (%)	0948	51.460	0.18000	51.874	0.61735	0.28276	190	-0.67	002.06	0
002	Protein (%)	0683	51.460	0.66000	51.874	0.61735	0.28276	190	-0.67	002.05	0
002	Protein (%)	0001	51.460	0.78000	51.874	0.61735	0.28276	190	-0.67	002.06	0
002	Protein (%)	0629	51.495	0.21000	51.874	0.61735	0.28276	190	-0.61	002.10	0
002	Protein (%)	0652	51.500	0.40000	51.874	0.61735	0.28276	190	-0.61	002.01	0
002	Protein (%)	0955	51.500	0.40000	51.874	0.61735	0.28276	190	-0.61	002.05	0
002	Protein (%)	0941	51.505	0.45000	51.874	0.61735	0.28276	190	-0.60	002.06	0
002	Protein (%)	0682	51.510	0.00000	51.874	0.61735	0.28276	190	-0.59	002.06	0
002	Protein (%)	0638	51.515	0.05000	51.874	0.61735	0.28276	190	-0.58	002.04	0
002	Protein (%)	0510	51.550	0.10000	51.874	0.61735	0.28276	190	-0.52	002.06	0
002	Protein (%)	0202	51.560	0.46000	51.874	0.61735	0.28276	190	-0.51	002.06	0
002	Protein (%)	0529	51.580	0.16000	51.874	0.61735	0.28276	190	-0.48	002.06	0
002	Protein (%)	0028	51.585	0.03000	51.874	0.61735	0.28276	190	-0.47	002.01	0
002	Protein (%)	0726	51.585	0.03000	51.874	0.61735	0.28276	190	-0.47	002.06	0
002	Protein (%)	0940	51.590	0.00000	51.874	0.61735	0.28276	190	-0.46	002.00	0
002	Protein (%)	0895	51.625	0.01000	51.874	0.61735	0.28276	190	-0.40	002.05	0
002	Protein (%)	0939	51.635	0.35000	51.874	0.61735	0.28276	190	-0.39	002.05	0
002	Protein (%)	0619	51.650	0.10000	51.874	0.61735	0.28276	190	-0.36	002.06	0
002	Protein (%)	0956	51.650	0.90000	51.874	0.61735	0.28276	190	-0.36	002.06	0
002	Protein (%)	0034	51.689	0.24600	51.874	0.61735	0.28276	190	-0.30	002.06	0
002	Protein (%)	0294	51.700	0.20000	51.874	0.61735	0.28276	190	-0.28	002.06	0
002	Protein (%)	0504	51.715	1.0700	51.874	0.61735	0.28276	190	-0.26	002.04	0
002	Protein (%)	0358	51.715	0.93000	51.874	0.61735	0.28276	190	-0.26	002.06	0
002	Protein (%)	0589	51.715	0.01000	51.874	0.61735	0.28276	190	-0.26	002.06	0
002	Protein (%)	0878	51.725	0.01000	51.874	0.61735	0.28276	190	-0.24	002.05	0
002	Protein (%)	0152	51.737	0.24500	51.874	0.61735	0.28276	190	-0.22	002.01	0
002	Protein (%)	0948	51.745	0.03000	51.874	0.61735	0.28276	190	-0.21	002.05	0
002	Protein (%)	0910	51.750	0.30000	51.874	0.61735	0.28276	190	-0.20	002.06	0
002	Protein (%)	0849	51.772	0.05600	51.874	0.61735	0.28276	190	-0.17	002.05	0
002	Protein (%)	0553	51.785	0.97000	51.874	0.61735	0.28276	190	-0.14	002.06	0
002	Protein (%)	0190	51.790	0.46000	51.874	0.61735	0.28276	190	-0.14	002.06	0
002	Protein (%)	0884	51.800	1.0000	51.874	0.61735	0.28276	190	-0.12	002.06	0
002	Protein (%)	0035	51.810	0.02000	51.874	0.61735	0.28276	190	-0.10	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein (%)	0100	51.825	0.33000	51.874	0.61735	0.28276	190	-0.08	002.06	0
002	Protein (%)	0683	51.825	0.01000	51.874	0.61735	0.28276	190	-0.08	002.06	0
002	Protein (%)	0300	51.840	0.22000	51.874	0.61735	0.28276	190	-0.05	002.06	0
002	Protein (%)	0878	51.845	0.21000	51.874	0.61735	0.28276	190	-0.05	002.06	0
002	Protein (%)	0541	51.870	0.24000	51.874	0.61735	0.28276	190	-0.01	002.06	0
002	Protein (%)	2059	51.875	0.54370	51.874	0.61735	0.28276	190	0.00	002.06	0
002	Protein (%)	0018	51.875	0.07000	51.874	0.61735	0.28276	190	0.00	002.06	0
002	Protein (%)	0309	51.879	0.09750	51.874	0.61735	0.28276	190	0.01	002.06	0
002	Protein (%)	0037	51.885	0.15000	51.874	0.61735	0.28276	190	0.02	002.06	0
002	Protein (%)	2053	51.895	0.29000	51.874	0.61735	0.28276	190	0.03	002.06	0
002	Protein (%)	0014	51.900	0.20000	51.874	0.61735	0.28276	190	0.04	002.06	0
002	Protein (%)	0903	51.915	0.01000	51.874	0.61735	0.28276	190	0.07	002.05	0
002	Protein (%)	0003	51.940	0.08000	51.874	0.61735	0.28276	190	0.11	002.06	0
002	Protein (%)	0413	51.950	0.50000	51.874	0.61735	0.28276	190	0.12	002.06	0
002	Protein (%)	0955	51.950	0.30000	51.874	0.61735	0.28276	190	0.12	002.06	0
002	Protein (%)	0505	51.960	0.00000	51.874	0.61735	0.28276	190	0.14	002.06	0
002	Protein (%)	0861	51.960	0.26000	51.874	0.61735	0.28276	190	0.14	002.06	0
002	Protein (%)	0650	51.970	0.44000	51.874	0.61735	0.28276	190	0.16	002.06	0
002	Protein (%)	0263	51.990	0.03360	51.874	0.61735	0.28276	190	0.19	002.06	0
002	Protein (%)	0148	51.990	0.08000	51.874	0.61735	0.28276	190	0.19	002.06	0
002	Protein (%)	0644	51.992	0.07200	51.874	0.61735	0.28276	190	0.19	002.06	0
002	Protein (%)	0205	51.995	0.03000	51.874	0.61735	0.28276	190	0.20	002.06	0
002	Protein (%)	0098	52.000	1.2000	51.874	0.61735	0.28276	190	0.20	002.06	0
002	Protein (%)	0626	52.012	0.00100	51.874	0.61735	0.28276	190	0.22	002.06	0
002	Protein (%)	0049	52.030	0.46000	51.874	0.61735	0.28276	190	0.25	002.06	0
002	Protein (%)	0171	52.035	0.15000	51.874	0.61735	0.28276	190	0.26	002.06	0
002	Protein (%)	2062	52.044	0.02910	51.874	0.61735	0.28276	190	0.28	002.05	0
002	Protein (%)	0036	52.045	0.19000	51.874	0.61735	0.28276	190	0.28	002.06	0
002	Protein (%)	0353	52.045	0.27000	51.874	0.61735	0.28276	190	0.28	002.06	0
002	Protein (%)	0051	52.050	0.30000	51.874	0.61735	0.28276	190	0.29	002.06	0
002	Protein (%)	0425	52.050	0.08000	51.874	0.61735	0.28276	190	0.29	002.06	0
002	Protein (%)	2004	52.050	0.10000	51.874	0.61735	0.28276	190	0.29	002.99	0
002	Protein (%)	0036	52.054	0.28280	51.874	0.61735	0.28276	190	0.29	002.02	0
002	Protein (%)	0618	52.058	0.03800	51.874	0.61735	0.28276	190	0.30	002.06	0
002	Protein (%)	0407	52.063	0.37100	51.874	0.61735	0.28276	190	0.31	002.06	0
002	Protein (%)	0574	52.065	0.07000	51.874	0.61735	0.28276	190	0.31	002.06	0
002	Protein (%)	0897	52.080	0.02000	51.874	0.61735	0.28276	190	0.33	002.05	0
002	Protein (%)	0571	52.085	0.04300	51.874	0.61735	0.28276	190	0.34	002.06	0
002	Protein (%)	2057	52.092	0.14400	51.874	0.61735	0.28276	190	0.35	002.06	0
002	Protein (%)	0646	52.095	0.01000	51.874	0.61735	0.28276	190	0.36	002.06	0
002	Protein (%)	0017	52.100	0.20000	51.874	0.61735	0.28276	190	0.37	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein (%)	0226	52.100	0.00000	51.874	0.61735	0.28276	190	0.37	002.06	0
002	Protein (%)	0610	52.100	0.00000	51.874	0.61735	0.28276	190	0.37	002.06	0
002	Protein (%)	0354	52.105	0.01000	51.874	0.61735	0.28276	190	0.37	002.06	0
002	Protein (%)	0843	52.110	0.22000	51.874	0.61735	0.28276	190	0.38	002.06	0
002	Protein (%)	0964	52.112	0.30600	51.874	0.61735	0.28276	190	0.39	002.06	0
002	Protein (%)	0233	52.115	0.27000	51.874	0.61735	0.28276	190	0.39	002.06	0
002	Protein (%)	2089	52.125	0.05000	51.874	0.61735	0.28276	190	0.41	002.06	0
002	Protein (%)	0676	52.130	0.50000	51.874	0.61735	0.28276	190	0.41	002.06	0
002	Protein (%)	0630	52.135	0.35000	51.874	0.61735	0.28276	190	0.42	002.06	0
002	Protein (%)	0010	52.150	0.10000	51.874	0.61735	0.28276	190	0.45	002.06	0
002	Protein (%)	0033	52.150	0.10000	51.874	0.61735	0.28276	190	0.45	002.06	0
002	Protein (%)	0520	52.150	0.10000	51.874	0.61735	0.28276	190	0.45	002.06	0
002	Protein (%)	0357	52.156	0.06250	51.874	0.61735	0.28276	190	0.46	002.06	0
002	Protein (%)	0089	52.160	0.00000	51.874	0.61735	0.28276	190	0.46	002.06	0
002	Protein (%)	0015	52.165	0.13000	51.874	0.61735	0.28276	190	0.47	002.06	0
002	Protein (%)	0687	52.175	0.15000	51.874	0.61735	0.28276	190	0.49	002.06	0
002	Protein (%)	0229	52.185	0.13000	51.874	0.61735	0.28276	190	0.50	002.06	0
002	Protein (%)	2155	52.192	0.88300	51.874	0.61735	0.28276	190	0.51	002.06	0
002	Protein (%)	0042	52.195	0.17000	51.874	0.61735	0.28276	190	0.52	002.02	0
002	Protein (%)	0508	52.196	0.04800	51.874	0.61735	0.28276	190	0.52	002.06	0
002	Protein (%)	0019	52.200	0.04000	51.874	0.61735	0.28276	190	0.53	002.06	0
002	Protein (%)	0175	52.200	0.20000	51.874	0.61735	0.28276	190	0.53	002.06	0
002	Protein (%)	0164	52.225	0.15000	51.874	0.61735	0.28276	190	0.57	002.06	0
002	Protein (%)	0278	52.250	0.50000	51.874	0.61735	0.28276	190	0.61	002.06	0
002	Protein (%)	0083	52.255	0.15000	51.874	0.61735	0.28276	190	0.62	002.06	0
002	Protein (%)	0693	52.256	0.04400	51.874	0.61735	0.28276	190	0.62	002.06	0
002	Protein (%)	0939	52.265	0.21000	51.874	0.61735	0.28276	190	0.63	002.06	0
002	Protein (%)	0581	52.275	0.45000	51.874	0.61735	0.28276	190	0.65	002.05	0
002	Protein (%)	0011	52.280	0.46000	51.874	0.61735	0.28276	190	0.66	002.11	0
002	Protein (%)	0504	52.295	0.31000	51.874	0.61735	0.28276	190	0.68	002.06	0
002	Protein (%)	0045	52.300	0.40000	51.874	0.61735	0.28276	190	0.69	002.06	0
002	Protein (%)	0142	52.305	0.01000	51.874	0.61735	0.28276	190	0.70	002.06	0
002	Protein (%)	0857	52.325	0.07000	51.874	0.61735	0.28276	190	0.73	002.06	0
002	Protein (%)	2076	52.347	0.11500	51.874	0.61735	0.28276	190	0.77	002.06	0
002	Protein (%)	0242	52.360	0.36000	51.874	0.61735	0.28276	190	0.79	002.06	0
002	Protein (%)	0139	52.365	0.17000	51.874	0.61735	0.28276	190	0.80	002.06	0
002	Protein (%)	0011	52.380	0.22000	51.874	0.61735	0.28276	190	0.82	002.06	0
002	Protein (%)	0298	52.390	0.02000	51.874	0.61735	0.28276	190	0.84	002.06	0
002	Protein (%)	0123	52.400	0.02000	51.874	0.61735	0.28276	190	0.85	002.06	0
002	Protein (%)	0186	52.400	0.00000	51.874	0.61735	0.28276	190	0.85	002.06	0
002	Protein (%)	0144	52.425	0.03000	51.874	0.61735	0.28276	190	0.89	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein (%)	0590	52.425	0.05000	51.874	0.61735	0.28276	190	0.89	002.06	0
002	Protein (%)	0692	52.450	0.50000	51.874	0.61735	0.28276	190	0.93	002.06	0
002	Protein (%)	0511	52.455	0.39000	51.874	0.61735	0.28276	190	0.94	002.06	0
002	Protein (%)	2033	52.455	0.11000	51.874	0.61735	0.28276	190	0.94	002.06	0
002	Protein (%)	0870	52.466	0.13400	51.874	0.61735	0.28276	190	0.96	002.01	0
002	Protein (%)	0366	52.500	0.40000	51.874	0.61735	0.28276	190	1.01	002.06	0
002	Protein (%)	0042	52.520	0.46000	51.874	0.61735	0.28276	190	1.05	002.06	0
002	Protein (%)	0512	52.520	0.04000	51.874	0.61735	0.28276	190	1.05	002.06	0
002	Protein (%)	0675	52.520	0.14000	51.874	0.61735	0.28276	190	1.05	002.06	0
002	Protein (%)	0047	52.550	0.70000	51.874	0.61735	0.28276	190	1.10	002.06	0
002	Protein (%)	0032	52.551	0.26900	51.874	0.61735	0.28276	190	1.10	002.06	0
002	Protein (%)	2129	52.610	1.1800	51.874	0.61735	0.28276	190	1.19	002.06	0
002	Protein (%)	0852	52.625	0.17000	51.874	0.61735	0.28276	190	1.22	002.05	0
002	Protein (%)	2012	52.650	0.02000	51.874	0.61735	0.28276	190	1.26	002.05	0
002	Protein (%)	2150	52.705	1.1900	51.874	0.61735	0.28276	190	1.35	002.06	0
002	Protein (%)	0848	52.750	0.08000	51.874	0.61735	0.28276	190	1.42	002.06	0
002	Protein (%)	0872	52.753	0.08200	51.874	0.61735	0.28276	190	1.42	002.06	0
002	Protein (%)	0889	52.805	0.25000	51.874	0.61735	0.28276	190	1.51	002.11	0
002	Protein (%)	0953	52.845	0.41000	51.874	0.61735	0.28276	190	1.57	002.06	0
002	Protein (%)	0208	52.850	0.90000	51.874	0.61735	0.28276	190	1.58	002.06	0
002	Protein (%)	0345	52.850	0.90000	51.874	0.61735	0.28276	190	1.58	002.06	0
002	Protein (%)	0870	52.917	1.8000	51.874	0.61735	0.28276	190	1.69	002.06	0
002	Protein (%)	0563	53.012	0.14580	51.874	0.61735	0.28276	190	1.84	002.08	0
002	Protein (%)	0685	53.120	0.26000	51.874	0.61735	0.28276	190	2.02	002.01	0
002	Protein (%)	0880	53.150	1.0800	51.874	0.61735	0.28276	190	2.07	002.06	0
002	Protein (%)	0876	53.215	0.03000	51.874	0.61735	0.28276	190	2.17	002.06	0
002	Protein (%)	0592	53.430	0.08000	51.874	0.61735	0.28276	190	2.52	002.06	0
002	Protein (%)	0539	53.710	0.00000	51.874	0.61735	0.28276	190	2.97	002.06	0
002	Protein (%)	0615	54.725	0.13000	51.874	0.61735	0.28276	190	4.62	002.06	0
002	Protein (%)	2012	55.210	0.04000	51.874	0.61735	0.28276	190	5.40	002.06	0
002	Protein (%)	2105	248.77	400.05	51.874	0.61735	0.28276	190	318.93	002.11	2
003	Fat (%)	2105	13.415	0.13000	18.059	0.35648	0.21946	139	-13.03	003.11	0
003	Fat (%)	0852	14.515	0.21000	18.059	0.35648	0.21946	139	-9.94	003.11	0
003	Fat (%)	0278	15.350	0.30000	18.059	0.35648	0.21946	139	-7.60	003.14	0
003	Fat (%)	0297	15.675	0.73000	18.059	0.35648	0.21946	139	-6.69	003.11	0
003	Fat (%)	0683	16.980	0.08000	18.059	0.35648	0.21946	139	-3.03	003.14	0
003	Fat (%)	0843	17.045	1.5500	18.059	0.35648	0.21946	139	-2.84	003.14	0
003	Fat (%)	0693	17.142	0.94000	18.059	0.35648	0.21946	139	-2.57	003.10	0
003	Fat (%)	0638	17.350	0.70000	18.059	0.35648	0.21946	139	-1.99	003.09	0
003	Fat (%)	0042	17.460	0.70000	18.059	0.35648	0.21946	139	-1.68	003.10	0
003	Fat (%)	0910	17.490	1.4000	18.059	0.35648	0.21946	139	-1.60	003.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat (%)	0357	17.550	0.10000	18.059	0.35648	0.21946	139	-1.43	003.14	0
003	Fat (%)	0646	17.595	0.75000	18.059	0.35648	0.21946	139	-1.30	003.13	0
003	Fat (%)	0590	17.600	0.30000	18.059	0.35648	0.21946	139	-1.29	003.09	0
003	Fat (%)	0912	17.600	0.30000	18.059	0.35648	0.21946	139	-1.29	003.14	0
003	Fat (%)	0051	17.650	0.10000	18.059	0.35648	0.21946	139	-1.15	003.10	0
003	Fat (%)	0861	17.655	0.27000	18.059	0.35648	0.21946	139	-1.13	003.10	0
003	Fat (%)	0728	17.660	0.02000	18.059	0.35648	0.21946	139	-1.12	003.10	0
003	Fat (%)	0895	17.665	0.01000	18.059	0.35648	0.21946	139	-1.10	003.12	0
003	Fat (%)	2129	17.675	0.15000	18.059	0.35648	0.21946	139	-1.08	003.14	0
003	Fat (%)	0884	17.700	0.40000	18.059	0.35648	0.21946	139	-1.01	003.06	0
003	Fat (%)	0366	17.700	0.40000	18.059	0.35648	0.21946	139	-1.01	003.10	0
003	Fat (%)	0619	17.700	0.20000	18.059	0.35648	0.21946	139	-1.01	003.10	0
003	Fat (%)	0036	17.703	0.04640	18.059	0.35648	0.21946	139	-1.00	003.10	0
003	Fat (%)	0623	17.735	0.59500	18.059	0.35648	0.21946	139	-0.91	003.10	0
003	Fat (%)	0263	17.740	0.15360	18.059	0.35648	0.21946	139	-0.89	003.09	0
003	Fat (%)	0619	17.750	0.10000	18.059	0.35648	0.21946	139	-0.87	003.14	0
003	Fat (%)	0948	17.770	0.04000	18.059	0.35648	0.21946	139	-0.81	003.09	0
003	Fat (%)	0553	17.770	0.36000	18.059	0.35648	0.21946	139	-0.81	003.10	0
003	Fat (%)	2089	17.780	0.36000	18.059	0.35648	0.21946	139	-0.78	003.10	0
003	Fat (%)	0581	17.780	0.10000	18.059	0.35648	0.21946	139	-0.78	003.14	0
003	Fat (%)	0512	17.790	0.88000	18.059	0.35648	0.21946	139	-0.75	003.09	0
003	Fat (%)	2073	17.795	0.29000	18.059	0.35648	0.21946	139	-0.74	003.10	0
003	Fat (%)	0294	17.820	0.14000	18.059	0.35648	0.21946	139	-0.67	003.06	0
003	Fat (%)	0298	17.820	0.02000	18.059	0.35648	0.21946	139	-0.67	003.10	0
003	Fat (%)	0144	17.820	0.04000	18.059	0.35648	0.21946	139	-0.67	003.14	0
003	Fat (%)	0045	17.850	0.10000	18.059	0.35648	0.21946	139	-0.59	003.10	0
003	Fat (%)	2144	17.850	0.04000	18.059	0.35648	0.21946	139	-0.59	003.99	0
003	Fat (%)	0675	17.880	0.08000	18.059	0.35648	0.21946	139	-0.50	003.09	0
003	Fat (%)	0297	17.885	0.07000	18.059	0.35648	0.21946	139	-0.49	003.06	0
003	Fat (%)	0674	17.890	0.00000	18.059	0.35648	0.21946	139	-0.47	003.09	0
003	Fat (%)	0345	17.900	0.40000	18.059	0.35648	0.21946	139	-0.45	003.00	0
003	Fat (%)	0226	17.900	0.00000	18.059	0.35648	0.21946	139	-0.45	003.09	0
003	Fat (%)	0186	17.900	0.20000	18.059	0.35648	0.21946	139	-0.45	003.14	0
003	Fat (%)	0520	17.900	0.04000	18.059	0.35648	0.21946	139	-0.45	003.14	0
003	Fat (%)	0956	17.900	0.20000	18.059	0.35648	0.21946	139	-0.45	003.14	0
003	Fat (%)	0037	17.905	0.01000	18.059	0.35648	0.21946	139	-0.43	003.14	0
003	Fat (%)	0948	17.905	0.01000	18.059	0.35648	0.21946	139	-0.43	003.14	0
003	Fat (%)	0353	17.910	0.18000	18.059	0.35648	0.21946	139	-0.42	003.09	0
003	Fat (%)	0066	17.950	0.20000	18.059	0.35648	0.21946	139	-0.30	003.14	0
003	Fat (%)	0598	17.955	0.11000	18.059	0.35648	0.21946	139	-0.29	003.14	0
003	Fat (%)	0511	17.980	0.30000	18.059	0.35648	0.21946	139	-0.22	003.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat (%)	0903	17.980	0.02000	18.059	0.35648	0.21946	139	-0.22	003.12	0
003	Fat (%)	0098	17.980	0.06000	18.059	0.35648	0.21946	139	-0.22	003.09	0
003	Fat (%)	0870	17.981	0.11100	18.059	0.35648	0.21946	139	-0.22	003.10	0
003	Fat (%)	0689	18.000	0.20000	18.059	0.35648	0.21946	139	-0.16	003.06	0
003	Fat (%)	0208	18.000	0.20000	18.059	0.35648	0.21946	139	-0.16	003.13	0
003	Fat (%)	0108	18.005	0.11000	18.059	0.35648	0.21946	139	-0.15	003.14	0
003	Fat (%)	0682	18.010	0.00000	18.059	0.35648	0.21946	139	-0.14	003.14	0
003	Fat (%)	0574	18.015	0.01000	18.059	0.35648	0.21946	139	-0.12	003.14	0
003	Fat (%)	0563	18.015	0.31870	18.059	0.35648	0.21946	139	-0.12	003.06	0
003	Fat (%)	0653	18.020	0.06000	18.059	0.35648	0.21946	139	-0.11	003.00	0
003	Fat (%)	0233	18.020	0.10000	18.059	0.35648	0.21946	139	-0.11	003.10	0
003	Fat (%)	0504	18.020	0.16000	18.059	0.35648	0.21946	139	-0.11	003.01	0
003	Fat (%)	0229	18.025	0.15000	18.059	0.35648	0.21946	139	-0.09	003.06	0
003	Fat (%)	0968	18.026	0.14100	18.059	0.35648	0.21946	139	-0.09	003.14	0
003	Fat (%)	0897	18.035	0.01000	18.059	0.35648	0.21946	139	-0.07	003.10	0
003	Fat (%)	0202	18.035	0.11000	18.059	0.35648	0.21946	139	-0.07	003.14	0
003	Fat (%)	2068	18.040	0.00000	18.059	0.35648	0.21946	139	-0.05	003.09	0
003	Fat (%)	0098	18.040	0.30000	18.059	0.35648	0.21946	139	-0.05	003.10	0
003	Fat (%)	0878	18.045	0.07000	18.059	0.35648	0.21946	139	-0.04	003.10	0
003	Fat (%)	0720	18.050	0.02000	18.059	0.35648	0.21946	139	-0.02	003.10	0
003	Fat (%)	0955	18.050	0.30000	18.059	0.35648	0.21946	139	-0.02	003.14	0
003	Fat (%)	0529	18.050	0.28000	18.059	0.35648	0.21946	139	-0.02	003.14	0
003	Fat (%)	0692	18.050	0.50000	18.059	0.35648	0.21946	139	-0.02	003.14	0
003	Fat (%)	0630	18.065	0.59000	18.059	0.35648	0.21946	139	0.02	003.09	0
003	Fat (%)	0100	18.070	0.18000	18.059	0.35648	0.21946	139	0.03	003.10	0
003	Fat (%)	2022	18.070	0.28000	18.059	0.35648	0.21946	139	0.03	003.10	0
003	Fat (%)	0407	18.083	0.25260	18.059	0.35648	0.21946	139	0.07	003.14	0
003	Fat (%)	0618	18.084	0.12800	18.059	0.35648	0.21946	139	0.07	003.10	0
003	Fat (%)	0644	18.086	0.00100	18.059	0.35648	0.21946	139	0.08	003.06	0
003	Fat (%)	0722	18.089	0.00000	18.059	0.35648	0.21946	139	0.09	003.09	0
003	Fat (%)	0242	18.090	0.02000	18.059	0.35648	0.21946	139	0.09	003.10	0
003	Fat (%)	0676	18.095	0.13000	18.059	0.35648	0.21946	139	0.10	003.10	0
003	Fat (%)	0123	18.100	0.20000	18.059	0.35648	0.21946	139	0.12	003.13	0
003	Fat (%)	0229	18.100	0.00000	18.059	0.35648	0.21946	139	0.12	003.14	0
003	Fat (%)	0027	18.101	0.01800	18.059	0.35648	0.21946	139	0.12	003.09	0
003	Fat (%)	0098	18.110	0.22000	18.059	0.35648	0.21946	139	0.14	003.13	0
003	Fat (%)	0536	18.110	0.16000	18.059	0.35648	0.21946	139	0.14	003.99	0
003	Fat (%)	0940	18.115	0.33000	18.059	0.35648	0.21946	139	0.16	003.12	0
003	Fat (%)	0033	18.115	0.37000	18.059	0.35648	0.21946	139	0.16	003.13	0
003	Fat (%)	0849	18.126	0.08800	18.059	0.35648	0.21946	139	0.19	003.06	0
003	Fat (%)	2006	18.160	0.08000	18.059	0.35648	0.21946	139	0.28	003.13	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat (%)	0142	18.170	0.06000	18.059	0.35648	0.21946	139	0.31	003.00	0
003	Fat (%)	0300	18.175	0.03000	18.059	0.35648	0.21946	139	0.33	003.06	0
003	Fat (%)	0615	18.185	0.33000	18.059	0.35648	0.21946	139	0.35	003.00	0
003	Fat (%)	0354	18.185	0.01000	18.059	0.35648	0.21946	139	0.35	003.09	0
003	Fat (%)	0190	18.190	0.06000	18.059	0.35648	0.21946	139	0.37	003.14	0
003	Fat (%)	0164	18.195	0.07000	18.059	0.35648	0.21946	139	0.38	003.06	0
003	Fat (%)	2012	18.210	0.00000	18.059	0.35648	0.21946	139	0.42	003.09	0
003	Fat (%)	0148	18.225	0.03000	18.059	0.35648	0.21946	139	0.47	003.06	0
003	Fat (%)	0089	18.230	0.02000	18.059	0.35648	0.21946	139	0.48	003.14	0
003	Fat (%)	0425	18.240	0.06000	18.059	0.35648	0.21946	139	0.51	003.06	0
003	Fat (%)	2033	18.250	0.12000	18.059	0.35648	0.21946	139	0.54	003.00	0
003	Fat (%)	0687	18.250	0.10000	18.059	0.35648	0.21946	139	0.54	003.06	0
003	Fat (%)	0726	18.255	0.01000	18.059	0.35648	0.21946	139	0.55	003.14	0
003	Fat (%)	0682	18.260	0.00000	18.059	0.35648	0.21946	139	0.56	003.06	0
003	Fat (%)	0083	18.275	0.15000	18.059	0.35648	0.21946	139	0.61	003.06	0
003	Fat (%)	0552	18.280	0.10000	18.059	0.35648	0.21946	139	0.62	003.06	0
003	Fat (%)	0205	18.280	0.20000	18.059	0.35648	0.21946	139	0.62	003.13	0
003	Fat (%)	2062	18.289	0.00170	18.059	0.35648	0.21946	139	0.65	003.12	0
003	Fat (%)	0358	18.300	0.08000	18.059	0.35648	0.21946	139	0.68	003.14	0
003	Fat (%)	0001	18.302	0.18000	18.059	0.35648	0.21946	139	0.68	003.14	0
003	Fat (%)	0175	18.335	0.13000	18.059	0.35648	0.21946	139	0.78	003.00	0
003	Fat (%)	0049	18.335	0.27000	18.059	0.35648	0.21946	139	0.78	003.14	0
003	Fat (%)	2155	18.374	0.87400	18.059	0.35648	0.21946	139	0.88	003.10	0
003	Fat (%)	0171	18.420	0.22000	18.059	0.35648	0.21946	139	1.01	003.12	0
003	Fat (%)	0152	18.450	0.50000	18.059	0.35648	0.21946	139	1.10	003.00	0
003	Fat (%)	0541	18.480	0.04000	18.059	0.35648	0.21946	139	1.18	003.10	0
003	Fat (%)	0848	18.500	0.02000	18.059	0.35648	0.21946	139	1.24	003.14	0
003	Fat (%)	0885	18.544	0.13020	18.059	0.35648	0.21946	139	1.36	003.00	0
003	Fat (%)	0893	18.550	0.00000	18.059	0.35648	0.21946	139	1.38	003.12	0
003	Fat (%)	0413	18.550	0.50000	18.059	0.35648	0.21946	139	1.38	003.14	0
003	Fat (%)	0309	18.597	0.13860	18.059	0.35648	0.21946	139	1.51	003.00	0
003	Fat (%)	0685	18.625	0.05000	18.059	0.35648	0.21946	139	1.59	003.09	0
003	Fat (%)	0032	18.635	0.87000	18.059	0.35648	0.21946	139	1.62	003.00	0
003	Fat (%)	0918	18.670	0.62000	18.059	0.35648	0.21946	139	1.71	003.06	0
003	Fat (%)	0876	18.675	0.25000	18.059	0.35648	0.21946	139	1.73	003.14	0
003	Fat (%)	0651	18.713	0.11000	18.059	0.35648	0.21946	139	1.84	003.10	0
003	Fat (%)	0265	18.835	0.27000	18.059	0.35648	0.21946	139	2.18	003.14	0
003	Fat (%)	0546	19.155	0.11000	18.059	0.35648	0.21946	139	3.08	003.99	0
003	Fat (%)	0941	19.270	0.04000	18.059	0.35648	0.21946	139	3.40	003.99	0
003	Fat (%)	0183	19.509	0.11300	18.059	0.35648	0.21946	139	4.07	003.14	0
003	Fat (%)	2009	19.589	0.00430	18.059	0.35648	0.21946	139	4.29	003.10	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat (%)	0004	19.765	0.11000	18.059	0.35648	0.21946	139	4.79	003.09	0
003	Fat (%)	0720	19.950	0.04000	18.059	0.35648	0.21946	139	5.31	003.11	0
003	Fat (%)	0889	20.110	0.84000	18.059	0.35648	0.21946	139	5.75	003.11	0
003	Fat (%)	0914	20.220	1.1000	18.059	0.35648	0.21946	139	6.06	003.99	0
003	Fat (%)	2145	21.205	0.13000	18.059	0.35648	0.21946	139	8.83	003.99	0
003	Fat (%)	0011	21.705	0.33000	18.059	0.35648	0.21946	139	10.23	003.11	0
003	Fat (%)	0880	12.495	5.9900	18.059	0.35648	0.21946	139	-15.61	003.99	1
004	Fiber (%)	0708	0.26000	0.10000	1.2401	0.36876	0.13964	110	-2.66	004.07	0
004	Fiber (%)	0265	0.43500	0.01000	1.2401	0.36876	0.13964	110	-2.18	004.07	0
004	Fiber (%)	0939	0.46500	0.19000	1.2401	0.36876	0.13964	110	-2.10	004.07	0
004	Fiber (%)	2009	0.49945	0.01990	1.2401	0.36876	0.13964	110	-2.01	004.07	0
004	Fiber (%)	0619	0.61100	0.14200	1.2401	0.36876	0.13964	110	-1.71	004.07	0
004	Fiber (%)	0169	0.61500	0.03000	1.2401	0.36876	0.13964	110	-1.70	004.00	0
004	Fiber (%)	0366	0.70000	0.00000	1.2401	0.36876	0.13964	110	-1.46	004.01	0
004	Fiber (%)	0144	0.75000	0.18000	1.2401	0.36876	0.13964	110	-1.33	004.07	0
004	Fiber (%)	2145	0.76500	0.05000	1.2401	0.36876	0.13964	110	-1.29	004.99	0
004	Fiber (%)	0354	0.77500	0.01000	1.2401	0.36876	0.13964	110	-1.26	004.07	0
004	Fiber (%)	0242	0.80500	0.15000	1.2401	0.36876	0.13964	110	-1.18	004.07	0
004	Fiber (%)	0968	0.80900	0.02000	1.2401	0.36876	0.13964	110	-1.17	004.07	0
004	Fiber (%)	0552	0.86000	0.14000	1.2401	0.36876	0.13964	110	-1.03	004.07	0
004	Fiber (%)	0876	0.88000	0.02000	1.2401	0.36876	0.13964	110	-0.98	004.07	0
004	Fiber (%)	0529	0.94000	0.00000	1.2401	0.36876	0.13964	110	-0.81	004.07	0
004	Fiber (%)	0910	0.95000	0.10000	1.2401	0.36876	0.13964	110	-0.79	004.07	0
004	Fiber (%)	0345	0.95500	0.01000	1.2401	0.36876	0.13964	110	-0.77	004.00	0
004	Fiber (%)	2022	0.95500	0.03000	1.2401	0.36876	0.13964	110	-0.77	004.06	0
004	Fiber (%)	0504	0.96500	0.05000	1.2401	0.36876	0.13964	110	-0.75	004.00	0
004	Fiber (%)	0720	0.97000	0.18000	1.2401	0.36876	0.13964	110	-0.73	004.06	0
004	Fiber (%)	0941	0.97000	0.18000	1.2401	0.36876	0.13964	110	-0.73	004.06	0
004	Fiber (%)	2004	0.98800	0.02000	1.2401	0.36876	0.13964	110	-0.68	004.00	0
004	Fiber (%)	0164	1.0000	0.00000	1.2401	0.36876	0.13964	110	-0.65	004.00	0
004	Fiber (%)	0581	1.0000	0.20000	1.2401	0.36876	0.13964	110	-0.65	004.07	0
004	Fiber (%)	0505	1.0050	0.19000	1.2401	0.36876	0.13964	110	-0.64	004.07	0
004	Fiber (%)	0353	1.0100	0.08000	1.2401	0.36876	0.13964	110	-0.62	004.03	0
004	Fiber (%)	0297	1.0100	0.06000	1.2401	0.36876	0.13964	110	-0.62	004.07	0
004	Fiber (%)	0003	1.0150	0.25000	1.2401	0.36876	0.13964	110	-0.61	004.07	0
004	Fiber (%)	0298	1.0200	0.02000	1.2401	0.36876	0.13964	110	-0.60	004.00	0
004	Fiber (%)	0878	1.0350	0.01000	1.2401	0.36876	0.13964	110	-0.56	004.06	0
004	Fiber (%)	0885	1.0426	0.06290	1.2401	0.36876	0.13964	110	-0.54	004.06	0
004	Fiber (%)	0940	1.0450	0.05000	1.2401	0.36876	0.13964	110	-0.53	004.99	0
004	Fiber (%)	0042	1.0500	0.00000	1.2401	0.36876	0.13964	110	-0.52	004.07	0
004	Fiber (%)	0100	1.0500	0.02000	1.2401	0.36876	0.13964	110	-0.52	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber (%)	2053	1.0500	0.04000	1.2401	0.36876	0.13964	110	-0.52	004.07	0
004	Fiber (%)	0918	1.0600	0.08000	1.2401	0.36876	0.13964	110	-0.49	004.06	0
004	Fiber (%)	0563	1.0656	0.26340	1.2401	0.36876	0.13964	110	-0.47	004.07	0
004	Fiber (%)	2068	1.0700	0.00000	1.2401	0.36876	0.13964	110	-0.46	004.06	0
004	Fiber (%)	0033	1.0800	0.10000	1.2401	0.36876	0.13964	110	-0.43	004.07	0
004	Fiber (%)	0536	1.0950	0.23000	1.2401	0.36876	0.13964	110	-0.39	004.07	0
004	Fiber (%)	0425	1.1000	0.00000	1.2401	0.36876	0.13964	110	-0.38	004.00	0
004	Fiber (%)	2033	1.1000	0.20000	1.2401	0.36876	0.13964	110	-0.38	004.03	0
004	Fiber (%)	2006	1.1000	0.10000	1.2401	0.36876	0.13964	110	-0.38	004.06	0
004	Fiber (%)	0413	1.1000	0.00000	1.2401	0.36876	0.13964	110	-0.38	004.07	0
004	Fiber (%)	0682	1.1000	0.00000	1.2401	0.36876	0.13964	110	-0.38	004.07	0
004	Fiber (%)	0726	1.1000	0.00000	1.2401	0.36876	0.13964	110	-0.38	004.07	0
004	Fiber (%)	0171	1.1100	0.06000	1.2401	0.36876	0.13964	110	-0.35	004.00	0
004	Fiber (%)	0968	1.1145	0.04300	1.2401	0.36876	0.13964	110	-0.34	004.06	0
004	Fiber (%)	0646	1.1200	0.02000	1.2401	0.36876	0.13964	110	-0.33	004.07	0
004	Fiber (%)	0948	1.1300	0.02000	1.2401	0.36876	0.13964	110	-0.30	004.06	0
004	Fiber (%)	0309	1.1303	0.00180	1.2401	0.36876	0.13964	110	-0.30	004.00	0
004	Fiber (%)	0644	1.1345	0.01700	1.2401	0.36876	0.13964	110	-0.29	004.07	0
004	Fiber (%)	0626	1.1600	0.28000	1.2401	0.36876	0.13964	110	-0.22	004.07	0
004	Fiber (%)	2149	1.1665	0.28700	1.2401	0.36876	0.13964	110	-0.20	004.06	0
004	Fiber (%)	0861	1.1800	0.06000	1.2401	0.36876	0.13964	110	-0.16	004.07	0
004	Fiber (%)	2076	1.1811	0.11990	1.2401	0.36876	0.13964	110	-0.16	004.00	0
004	Fiber (%)	0208	1.1900	0.12000	1.2401	0.36876	0.13964	110	-0.14	004.00	0
004	Fiber (%)	0884	1.2000	0.00000	1.2401	0.36876	0.13964	110	-0.11	004.00	0
004	Fiber (%)	0692	1.2000	0.20000	1.2401	0.36876	0.13964	110	-0.11	004.07	0
004	Fiber (%)	0893	1.2100	0.00000	1.2401	0.36876	0.13964	110	-0.08	004.06	0
004	Fiber (%)	0693	1.2370	0.33200	1.2401	0.36876	0.13964	110	-0.01	004.03	0
004	Fiber (%)	0512	1.2485	0.24500	1.2401	0.36876	0.13964	110	0.02	004.06	0
004	Fiber (%)	0190	1.2500	0.10000	1.2401	0.36876	0.13964	110	0.03	004.07	0
004	Fiber (%)	2073	1.2550	0.11000	1.2401	0.36876	0.13964	110	0.04	004.00	0
004	Fiber (%)	0895	1.2550	0.01000	1.2401	0.36876	0.13964	110	0.04	004.06	0
004	Fiber (%)	0948	1.2550	0.07000	1.2401	0.36876	0.13964	110	0.04	004.07	0
004	Fiber (%)	0683	1.2750	0.27000	1.2401	0.36876	0.13964	110	0.09	004.07	0
004	Fiber (%)	0590	1.2800	0.38000	1.2401	0.36876	0.13964	110	0.11	004.06	0
004	Fiber (%)	0175	1.3050	0.03000	1.2401	0.36876	0.13964	110	0.18	004.00	0
004	Fiber (%)	0903	1.3150	0.01000	1.2401	0.36876	0.13964	110	0.20	004.06	0
004	Fiber (%)	0897	1.3250	0.03000	1.2401	0.36876	0.13964	110	0.23	004.06	0
004	Fiber (%)	0205	1.3450	0.13000	1.2401	0.36876	0.13964	110	0.28	004.06	0
004	Fiber (%)	0511	1.3500	0.50000	1.2401	0.36876	0.13964	110	0.30	004.00	0
004	Fiber (%)	2089	1.3600	0.36000	1.2401	0.36876	0.13964	110	0.33	004.03	0
004	Fiber (%)	2062	1.3635	0.00100	1.2401	0.36876	0.13964	110	0.33	004.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber (%)	0653	1.3700	0.06000	1.2401	0.36876	0.13964	110	0.35	004.07	0
004	Fiber (%)	0955	1.3950	0.07000	1.2401	0.36876	0.13964	110	0.42	004.07	0
004	Fiber (%)	0638	1.4000	0.00000	1.2401	0.36876	0.13964	110	0.43	004.06	0
004	Fiber (%)	0123	1.4100	0.06000	1.2401	0.36876	0.13964	110	0.46	004.07	0
004	Fiber (%)	0066	1.4150	0.09000	1.2401	0.36876	0.13964	110	0.47	004.07	0
004	Fiber (%)	0263	1.4266	0.04000	1.2401	0.36876	0.13964	110	0.51	004.07	0
004	Fiber (%)	0027	1.4400	0.00600	1.2401	0.36876	0.13964	110	0.54	004.06	0
004	Fiber (%)	0956	1.4500	0.50000	1.2401	0.36876	0.13964	110	0.57	004.07	0
004	Fiber (%)	0870	1.4549	0.04030	1.2401	0.36876	0.13964	110	0.58	004.07	0
004	Fiber (%)	0674	1.5000	0.00000	1.2401	0.36876	0.13964	110	0.70	004.06	0
004	Fiber (%)	0028	1.5000	0.20000	1.2401	0.36876	0.13964	110	0.70	004.07	0
004	Fiber (%)	0089	1.5100	0.02000	1.2401	0.36876	0.13964	110	0.73	004.07	0
004	Fiber (%)	0676	1.5300	0.02000	1.2401	0.36876	0.13964	110	0.79	004.06	0
004	Fiber (%)	0520	1.6000	0.10000	1.2401	0.36876	0.13964	110	0.98	004.07	0
004	Fiber (%)	0202	1.6350	0.43000	1.2401	0.36876	0.13964	110	1.07	004.07	0
004	Fiber (%)	0553	1.6450	0.15000	1.2401	0.36876	0.13964	110	1.10	004.07	0
004	Fiber (%)	0083	1.6500	0.20000	1.2401	0.36876	0.13964	110	1.11	004.07	0
004	Fiber (%)	0294	1.6500	0.10000	1.2401	0.36876	0.13964	110	1.11	004.07	0
004	Fiber (%)	0011	1.7000	0.66000	1.2401	0.36876	0.13964	110	1.25	004.11	0
004	Fiber (%)	0226	1.7500	0.10000	1.2401	0.36876	0.13964	110	1.38	004.00	0
004	Fiber (%)	2144	1.7700	0.06000	1.2401	0.36876	0.13964	110	1.44	004.99	0
004	Fiber (%)	0675	1.8050	0.09000	1.2401	0.36876	0.13964	110	1.53	004.06	0
004	Fiber (%)	0229	1.8050	0.01000	1.2401	0.36876	0.13964	110	1.53	004.07	0
004	Fiber (%)	0019	1.8800	0.06000	1.2401	0.36876	0.13964	110	1.74	004.07	0
004	Fiber (%)	0045	1.9000	0.00000	1.2401	0.36876	0.13964	110	1.79	004.03	0
004	Fiber (%)	0034	1.9100	0.64000	1.2401	0.36876	0.13964	110	1.82	004.07	0
004	Fiber (%)	0354	2.0500	0.12000	1.2401	0.36876	0.13964	110	2.20	004.00	0
004	Fiber (%)	0849	2.0920	0.01600	1.2401	0.36876	0.13964	110	2.31	004.07	0
004	Fiber (%)	0407	2.2604	0.09180	1.2401	0.36876	0.13964	110	2.77	004.07	0
004	Fiber (%)	0848	2.2900	0.78000	1.2401	0.36876	0.13964	110	2.85	004.07	0
004	Fiber (%)	0510	2.6000	0.60000	1.2401	0.36876	0.13964	110	3.69	004.07	0
004	Fiber (%)	0032	2.6600	1.3600	1.2401	0.36876	0.13964	110	3.85	004.07	0
004	Fiber (%)	0722	4.7307	0.01190	1.2401	0.36876	0.13964	110	9.47	004.06	0
004	Fiber (%)	2012	4.8600	0.28000	1.2401	0.36876	0.13964	110	9.82	004.07	0
004	Fiber (%)	0278	9.8000	1.0000	1.2401	0.36876	0.13964	110	23.21	004.07	0
004	Fiber (%)	0598	151.00	14.000	1.2401	0.36876	0.13964	110	406.12	004.07	2
004	Fiber (%)	0720	0.00000	0.00000	1.2401	0.36876	0.13964	110		004.11	4
005	Ash (%)	0297	9.3950	0.25000	10.833	0.13267	0.06667	143	-10.84	005.11	0
005	Ash (%)	0720	9.9700	0.16000	10.833	0.13267	0.06667	143	-6.51	005.11	0
005	Ash (%)	0874	10.385	0.13000	10.833	0.13267	0.06667	143	-3.38	005.00	0
005	Ash (%)	0546	10.390	0.04000	10.833	0.13267	0.06667	143	-3.34	005.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0674	10.410	0.00000	10.833	0.13267	0.06667	143	-3.19	005.00	0
005	Ash (%)	0169	10.540	0.04000	10.833	0.13267	0.06667	143	-2.21	005.00	0
005	Ash (%)	0278	10.555	0.01000	10.833	0.13267	0.06667	143	-2.10	005.00	0
005	Ash (%)	2033	10.600	0.20000	10.833	0.13267	0.06667	143	-1.76	005.00	0
005	Ash (%)	0539	10.600	0.06000	10.833	0.13267	0.06667	143	-1.76	005.00	0
005	Ash (%)	0358	10.615	0.15000	10.833	0.13267	0.06667	143	-1.65	005.00	0
005	Ash (%)	0098	10.620	0.02000	10.833	0.13267	0.06667	143	-1.61	005.00	0
005	Ash (%)	0581	10.655	0.27000	10.833	0.13267	0.06667	143	-1.34	005.00	0
005	Ash (%)	0144	10.655	0.05000	10.833	0.13267	0.06667	143	-1.34	005.00	0
005	Ash (%)	0589	10.665	0.27000	10.833	0.13267	0.06667	143	-1.27	005.00	0
005	Ash (%)	0643	10.675	0.01000	10.833	0.13267	0.06667	143	-1.19	005.00	0
005	Ash (%)	0574	10.675	0.11000	10.833	0.13267	0.06667	143	-1.19	005.99	0
005	Ash (%)	0876	10.685	0.01000	10.833	0.13267	0.06667	143	-1.12	005.00	0
005	Ash (%)	0541	10.685	0.07000	10.833	0.13267	0.06667	143	-1.12	005.00	0
005	Ash (%)	0893	10.690	0.00000	10.833	0.13267	0.06667	143	-1.08	005.05	0
005	Ash (%)	0004	10.690	0.04000	10.833	0.13267	0.06667	143	-1.08	005.00	0
005	Ash (%)	0651	10.695	0.02800	10.833	0.13267	0.06667	143	-1.04	005.00	0
005	Ash (%)	0019	10.700	0.00000	10.833	0.13267	0.06667	143	-1.01	005.00	0
005	Ash (%)	0047	10.700	0.00000	10.833	0.13267	0.06667	143	-1.01	005.00	0
005	Ash (%)	0692	10.700	0.00000	10.833	0.13267	0.06667	143	-1.01	005.05	0
005	Ash (%)	2012	10.715	0.01000	10.833	0.13267	0.06667	143	-0.89	005.00	0
005	Ash (%)	0693	10.716	0.11700	10.833	0.13267	0.06667	143	-0.89	005.00	0
005	Ash (%)	0066	10.725	0.07000	10.833	0.13267	0.06667	143	-0.82	005.05	0
005	Ash (%)	0049	10.725	0.15000	10.833	0.13267	0.06667	143	-0.82	005.00	0
005	Ash (%)	0164	10.725	0.01000	10.833	0.13267	0.06667	143	-0.82	005.00	0
005	Ash (%)	0175	10.730	0.02000	10.833	0.13267	0.06667	143	-0.78	005.00	0
005	Ash (%)	0001	10.732	0.00280	10.833	0.13267	0.06667	143	-0.77	005.00	0
005	Ash (%)	0142	10.735	0.03000	10.833	0.13267	0.06667	143	-0.74	005.00	0
005	Ash (%)	0520	10.735	0.03000	10.833	0.13267	0.06667	143	-0.74	005.00	0
005	Ash (%)	0354	10.740	0.04000	10.833	0.13267	0.06667	143	-0.70	005.00	0
005	Ash (%)	0425	10.745	0.05000	10.833	0.13267	0.06667	143	-0.67	005.00	0
005	Ash (%)	0208	10.750	0.10000	10.833	0.13267	0.06667	143	-0.63	005.00	0
005	Ash (%)	0148	10.755	0.01000	10.833	0.13267	0.06667	143	-0.59	005.00	0
005	Ash (%)	2068	10.755	0.07000	10.833	0.13267	0.06667	143	-0.59	005.00	0
005	Ash (%)	0940	10.755	0.11000	10.833	0.13267	0.06667	143	-0.59	005.05	0
005	Ash (%)	2089	10.760	0.02000	10.833	0.13267	0.06667	143	-0.55	005.00	0
005	Ash (%)	0941	10.760	0.08000	10.833	0.13267	0.06667	143	-0.55	005.05	0
005	Ash (%)	0035	10.765	0.05000	10.833	0.13267	0.06667	143	-0.52	005.00	0
005	Ash (%)	0229	10.765	0.21000	10.833	0.13267	0.06667	143	-0.52	005.00	0
005	Ash (%)	2073	10.765	0.07000	10.833	0.13267	0.06667	143	-0.52	005.00	0
005	Ash (%)	0644	10.770	0.06000	10.833	0.13267	0.06667	143	-0.48	005.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0675	10.775	0.03000	10.833	0.13267	0.06667	143	-0.44	005.00	0
005	Ash (%)	0897	10.775	0.01000	10.833	0.13267	0.06667	143	-0.44	005.05	0
005	Ash (%)	0650	10.775	0.13000	10.833	0.13267	0.06667	143	-0.44	005.00	0
005	Ash (%)	0123	10.780	0.18000	10.833	0.13267	0.06667	143	-0.40	005.00	0
005	Ash (%)	0682	10.780	0.00000	10.833	0.13267	0.06667	143	-0.40	005.00	0
005	Ash (%)	0870	10.783	0.00660	10.833	0.13267	0.06667	143	-0.38	005.00	0
005	Ash (%)	0726	10.785	0.01000	10.833	0.13267	0.06667	143	-0.36	005.05	0
005	Ash (%)	0205	10.790	0.06000	10.833	0.13267	0.06667	143	-0.33	005.00	0
005	Ash (%)	0511	10.790	0.06000	10.833	0.13267	0.06667	143	-0.33	005.00	0
005	Ash (%)	0512	10.790	0.08000	10.833	0.13267	0.06667	143	-0.33	005.00	0
005	Ash (%)	0171	10.795	0.03000	10.833	0.13267	0.06667	143	-0.29	005.00	0
005	Ash (%)	0505	10.795	0.01000	10.833	0.13267	0.06667	143	-0.29	005.00	0
005	Ash (%)	0510	10.795	0.07000	10.833	0.13267	0.06667	143	-0.29	005.00	0
005	Ash (%)	0598	10.795	0.01000	10.833	0.13267	0.06667	143	-0.29	005.00	0
005	Ash (%)	0848	10.795	0.01000	10.833	0.13267	0.06667	143	-0.29	005.00	0
005	Ash (%)	0051	10.800	0.00000	10.833	0.13267	0.06667	143	-0.25	005.00	0
005	Ash (%)	0242	10.800	0.16000	10.833	0.13267	0.06667	143	-0.25	005.00	0
005	Ash (%)	0956	10.800	0.00000	10.833	0.13267	0.06667	143	-0.25	005.00	0
005	Ash (%)	2006	10.800	0.02000	10.833	0.13267	0.06667	143	-0.25	005.00	0
005	Ash (%)	0413	10.800	0.00000	10.833	0.13267	0.06667	143	-0.25	005.05	0
005	Ash (%)	0263	10.805	0.01640	10.833	0.13267	0.06667	143	-0.22	005.05	0
005	Ash (%)	0504	10.805	0.01000	10.833	0.13267	0.06667	143	-0.21	005.00	0
005	Ash (%)	0653	10.805	0.03000	10.833	0.13267	0.06667	143	-0.21	005.00	0
005	Ash (%)	0083	10.815	0.17000	10.833	0.13267	0.06667	143	-0.14	005.00	0
005	Ash (%)	0536	10.830	0.08000	10.833	0.13267	0.06667	143	-0.03	005.99	0
005	Ash (%)	0529	10.830	0.10000	10.833	0.13267	0.06667	143	-0.03	005.00	0
005	Ash (%)	2022	10.830	0.04000	10.833	0.13267	0.06667	143	-0.03	005.00	0
005	Ash (%)	2053	10.835	0.03000	10.833	0.13267	0.06667	143	0.01	005.00	0
005	Ash (%)	2009	10.837	0.01750	10.833	0.13267	0.06667	143	0.03	005.05	0
005	Ash (%)	0861	10.845	0.11000	10.833	0.13267	0.06667	143	0.09	005.00	0
005	Ash (%)	0033	10.850	0.10000	10.833	0.13267	0.06667	143	0.13	005.00	0
005	Ash (%)	0045	10.850	0.10000	10.833	0.13267	0.06667	143	0.13	005.00	0
005	Ash (%)	0226	10.850	0.10000	10.833	0.13267	0.06667	143	0.13	005.00	0
005	Ash (%)	0912	10.850	0.10000	10.833	0.13267	0.06667	143	0.13	005.03	0
005	Ash (%)	0623	10.852	0.09300	10.833	0.13267	0.06667	143	0.14	005.00	0
005	Ash (%)	0903	10.855	0.01000	10.833	0.13267	0.06667	143	0.16	005.05	0
005	Ash (%)	0722	10.855	0.00970	10.833	0.13267	0.06667	143	0.16	005.00	0
005	Ash (%)	0849	10.856	0.07200	10.833	0.13267	0.06667	143	0.17	005.05	0
005	Ash (%)	0622	10.858	0.04670	10.833	0.13267	0.06667	143	0.18	005.00	0
005	Ash (%)	0689	10.865	0.01000	10.833	0.13267	0.06667	143	0.24	005.05	0
005	Ash (%)	0038	10.865	0.23000	10.833	0.13267	0.06667	143	0.24	005.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0407	10.867	0.01090	10.833	0.13267	0.06667	143	0.26	005.00	0
005	Ash (%)	0563	10.872	0.04700	10.833	0.13267	0.06667	143	0.29	005.00	0
005	Ash (%)	0139	10.875	0.03000	10.833	0.13267	0.06667	143	0.31	005.00	0
005	Ash (%)	2144	10.875	0.01000	10.833	0.13267	0.06667	143	0.31	005.99	0
005	Ash (%)	0015	10.880	0.12000	10.833	0.13267	0.06667	143	0.35	005.00	0
005	Ash (%)	0298	10.880	0.02000	10.833	0.13267	0.06667	143	0.35	005.00	0
005	Ash (%)	0953	10.880	0.24000	10.833	0.13267	0.06667	143	0.35	005.00	0
005	Ash (%)	0003	10.880	0.00000	10.833	0.13267	0.06667	143	0.35	005.05	0
005	Ash (%)	0968	10.887	0.02100	10.833	0.13267	0.06667	143	0.40	005.00	0
005	Ash (%)	0918	10.890	0.06000	10.833	0.13267	0.06667	143	0.43	005.00	0
005	Ash (%)	0553	10.895	0.09000	10.833	0.13267	0.06667	143	0.46	005.00	0
005	Ash (%)	2150	10.895	0.05000	10.833	0.13267	0.06667	143	0.46	005.00	0
005	Ash (%)	0683	10.900	0.22000	10.833	0.13267	0.06667	143	0.50	005.05	0
005	Ash (%)	0108	10.900	0.04000	10.833	0.13267	0.06667	143	0.50	005.00	0
005	Ash (%)	0357	10.900	0.00000	10.833	0.13267	0.06667	143	0.50	005.00	0
005	Ash (%)	0294	10.900	0.00000	10.833	0.13267	0.06667	143	0.50	005.05	0
005	Ash (%)	0652	10.900	0.00000	10.833	0.13267	0.06667	143	0.50	005.99	0
005	Ash (%)	0895	10.905	0.01000	10.833	0.13267	0.06667	143	0.54	005.05	0
005	Ash (%)	0152	10.915	0.11000	10.833	0.13267	0.06667	143	0.62	005.05	0
005	Ash (%)	0202	10.915	0.03000	10.833	0.13267	0.06667	143	0.62	005.99	0
005	Ash (%)	0309	10.920	0.10000	10.833	0.13267	0.06667	143	0.65	005.00	0
005	Ash (%)	0610	10.920	0.16000	10.833	0.13267	0.06667	143	0.65	005.05	0
005	Ash (%)	0353	10.925	0.05000	10.833	0.13267	0.06667	143	0.69	005.00	0
005	Ash (%)	0646	10.925	0.01000	10.833	0.13267	0.06667	143	0.69	005.00	0
005	Ash (%)	0878	10.925	0.07000	10.833	0.13267	0.06667	143	0.69	005.05	0
005	Ash (%)	0590	10.930	0.02000	10.833	0.13267	0.06667	143	0.73	005.05	0
005	Ash (%)	0885	10.931	0.07160	10.833	0.13267	0.06667	143	0.74	005.05	0
005	Ash (%)	0630	10.940	0.00000	10.833	0.13267	0.06667	143	0.80	005.00	0
005	Ash (%)	0638	10.940	0.04000	10.833	0.13267	0.06667	143	0.80	005.05	0
005	Ash (%)	0017	10.950	0.10000	10.833	0.13267	0.06667	143	0.88	005.00	0
005	Ash (%)	0186	10.950	0.10000	10.833	0.13267	0.06667	143	0.88	005.00	0
005	Ash (%)	0366	10.950	0.10000	10.833	0.13267	0.06667	143	0.88	005.00	0
005	Ash (%)	0884	10.950	0.10000	10.833	0.13267	0.06667	143	0.88	005.99	0
005	Ash (%)	0939	10.960	0.08000	10.833	0.13267	0.06667	143	0.95	005.05	0
005	Ash (%)	2149	10.975	0.01500	10.833	0.13267	0.06667	143	1.06	005.05	0
005	Ash (%)	0685	10.980	0.04000	10.833	0.13267	0.06667	143	1.11	005.05	0
005	Ash (%)	2062	10.980	0.00000	10.833	0.13267	0.06667	143	1.11	005.05	0
005	Ash (%)	2146	10.985	0.05000	10.833	0.13267	0.06667	143	1.14	005.05	0
005	Ash (%)	0629	10.985	0.05000	10.833	0.13267	0.06667	143	1.14	005.99	0
005	Ash (%)	2145	10.990	0.06000	10.833	0.13267	0.06667	143	1.18	005.99	0
005	Ash (%)	0100	11.000	0.06000	10.833	0.13267	0.06667	143	1.26	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0345	11.000	0.20000	10.833	0.13267	0.06667	143	1.26	005.00	0
005	Ash (%)	0619	11.000	0.00000	10.833	0.13267	0.06667	143	1.26	005.05	0
005	Ash (%)	0728	11.019	0.08300	10.833	0.13267	0.06667	143	1.40	005.99	0
005	Ash (%)	0615	11.030	0.02000	10.833	0.13267	0.06667	143	1.48	005.00	0
005	Ash (%)	0720	11.050	0.02000	10.833	0.13267	0.06667	143	1.63	005.05	0
005	Ash (%)	0618	11.084	0.13510	10.833	0.13267	0.06667	143	1.89	005.00	0
005	Ash (%)	0852	11.085	0.01000	10.833	0.13267	0.06667	143	1.90	005.00	0
005	Ash (%)	2129	11.085	0.17000	10.833	0.13267	0.06667	143	1.90	005.05	0
005	Ash (%)	0955	11.100	0.00000	10.833	0.13267	0.06667	143	2.01	005.05	0
005	Ash (%)	0676	11.115	0.01000	10.833	0.13267	0.06667	143	2.12	005.99	0
005	Ash (%)	0190	11.115	0.03000	10.833	0.13267	0.06667	143	2.12	005.05	0
005	Ash (%)	0910	11.150	0.10000	10.833	0.13267	0.06667	143	2.39	005.99	0
005	Ash (%)	0297	11.150	0.30000	10.833	0.13267	0.06667	143	2.39	005.00	0
005	Ash (%)	0948	11.195	0.01000	10.833	0.13267	0.06667	143	2.73	005.00	0
005	Ash (%)	0265	11.635	0.31000	10.833	0.13267	0.06667	143	6.04	005.00	0
005	Ash (%)	2105	13.885	0.17000	10.833	0.13267	0.06667	143	23.00	005.11	0
005	Ash (%)	0027	10.534	0.37700	10.833	0.13267	0.06667	143	-2.26	005.00	1
005	Ash (%)	0552	10.880	0.40000	10.833	0.13267	0.06667	143	0.35	005.00	1
005	Ash (%)	0880	11.060	0.60000	10.833	0.13267	0.06667	143	1.71	005.00	1
005	Ash (%)	0889	6.2800	0.10000	10.833	0.13267	0.06667	143	-34.32	005.11	2
005	Ash (%)	0852	15.115	0.07000	10.833	0.13267	0.06667	143	32.27	005.11	2
006	Total sugars (%)	0407	0.24000	0.00000				3		006.01	0
006	Total sugars (%)	0956	0.55000	0.10000				3		006.99	0
006	Total sugars (%)	0910	0.85000	0.10000				3		006.00	0
006	Total sugars (%)	0720	0.00000	0.00000				3		006.99	4
008	Fiber (%)	0720	0.54000	0.22000	3.6351	1.6521	0.22576	50	-1.87	008.99	0
008	Fiber (%)	0948	1.6300	0.04000	3.6351	1.6521	0.22576	50	-1.21	008.02	0
008	Fiber (%)	0910	1.8750	0.07000	3.6351	1.6521	0.22576	50	-1.07	008.08	0
008	Fiber (%)	0590	1.9000	0.20000	3.6351	1.6521	0.22576	50	-1.05	008.02	0
008	Fiber (%)	0956	1.9500	0.50000	3.6351	1.6521	0.22576	50	-1.02	008.08	0
008	Fiber (%)	0676	2.0200	0.10000	3.6351	1.6521	0.22576	50	-0.98	008.02	0
008	Fiber (%)	0870	2.1050	0.08570	3.6351	1.6521	0.22576	50	-0.93	008.08	0
008	Fiber (%)	0208	2.2200	0.04000	3.6351	1.6521	0.22576	50	-0.86	008.02	0
008	Fiber (%)	0861	2.2700	0.22000	3.6351	1.6521	0.22576	50	-0.83	008.08	0
008	Fiber (%)	0939	2.3450	0.37000	3.6351	1.6521	0.22576	50	-0.78	008.02	0
008	Fiber (%)	0066	2.3650	0.17000	3.6351	1.6521	0.22576	50	-0.77	008.08	0
008	Fiber (%)	0581	2.3650	0.11000	3.6351	1.6521	0.22576	50	-0.77	008.08	0
008	Fiber (%)	0353	2.4600	0.14000	3.6351	1.6521	0.22576	50	-0.71	008.02	0
008	Fiber (%)	0885	2.5045	0.02530	3.6351	1.6521	0.22576	50	-0.68	008.02	0
008	Fiber (%)	0358	2.5450	0.11000	3.6351	1.6521	0.22576	50	-0.66	008.08	0
008	Fiber (%)	0619	2.6200	0.44000	3.6351	1.6521	0.22576	50	-0.61	008.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
008	Fiber (%)	0297	2.6800	0.00000	3.6351	1.6521	0.22576	50	-0.58	008.08	0
008	Fiber (%)	0413	2.7000	0.20000	3.6351	1.6521	0.22576	50	-0.57	008.08	0
008	Fiber (%)	0148	2.7150	0.01000	3.6351	1.6521	0.22576	50	-0.56	008.02	0
008	Fiber (%)	0148	2.7600	0.66000	3.6351	1.6521	0.22576	50	-0.53	008.08	0
008	Fiber (%)	0309	2.8826	0.02110	3.6351	1.6521	0.22576	50	-0.46	008.02	0
008	Fiber (%)	0033	2.9150	0.03000	3.6351	1.6521	0.22576	50	-0.44	008.08	0
008	Fiber (%)	0918	3.0362	0.42850	3.6351	1.6521	0.22576	50	-0.36	008.08	0
008	Fiber (%)	0646	3.1450	0.19000	3.6351	1.6521	0.22576	50	-0.30	008.08	0
008	Fiber (%)	0354	3.1850	0.07000	3.6351	1.6521	0.22576	50	-0.27	008.08	0
008	Fiber (%)	0610	3.3500	0.30000	3.6351	1.6521	0.22576	50	-0.17	008.99	0
008	Fiber (%)	2004	3.3550	0.13000	3.6351	1.6521	0.22576	50	-0.17	008.08	0
008	Fiber (%)	0263	3.4338	0.02750	3.6351	1.6521	0.22576	50	-0.12	008.08	0
008	Fiber (%)	2129	3.6400	0.16000	3.6351	1.6521	0.22576	50	0.00	008.08	0
008	Fiber (%)	0049	3.6600	1.0000	3.6351	1.6521	0.22576	50	0.02	008.08	0
008	Fiber (%)	0190	3.7850	0.07000	3.6351	1.6521	0.22576	50	0.09	008.08	0
008	Fiber (%)	0726	3.8700	0.06000	3.6351	1.6521	0.22576	50	0.14	008.08	0
008	Fiber (%)	0226	4.1500	0.10000	3.6351	1.6521	0.22576	50	0.31	008.02	0
008	Fiber (%)	0529	4.2150	0.11000	3.6351	1.6521	0.22576	50	0.35	008.08	0
008	Fiber (%)	0004	4.3200	0.18000	3.6351	1.6521	0.22576	50	0.41	008.08	0
008	Fiber (%)	0001	4.6800	0.06000	3.6351	1.6521	0.22576	50	0.63	008.08	0
008	Fiber (%)	0278	4.9500	0.10000	3.6351	1.6521	0.22576	50	0.80	008.08	0
008	Fiber (%)	0357	5.0000	0.00000	3.6351	1.6521	0.22576	50	0.83	008.08	0
008	Fiber (%)	0096	5.1550	0.25000	3.6351	1.6521	0.22576	50	0.92	008.08	0
008	Fiber (%)	0037	5.1750	0.35000	3.6351	1.6521	0.22576	50	0.93	008.08	0
008	Fiber (%)	0164	5.5000	0.00000	3.6351	1.6521	0.22576	50	1.13	008.08	0
008	Fiber (%)	0675	5.6150	0.29000	3.6351	1.6521	0.22576	50	1.20	008.02	0
008	Fiber (%)	0045	5.6650	0.43000	3.6351	1.6521	0.22576	50	1.23	008.08	0
008	Fiber (%)	0693	5.6770	0.73600	3.6351	1.6521	0.22576	50	1.24	008.08	0
008	Fiber (%)	0083	5.8550	0.21000	3.6351	1.6521	0.22576	50	1.34	008.08	0
008	Fiber (%)	0265	5.9000	0.80000	3.6351	1.6521	0.22576	50	1.37	008.05	0
008	Fiber (%)	0536	6.1650	0.97000	3.6351	1.6521	0.22576	50	1.53	008.08	0
008	Fiber (%)	0294	6.3500	0.10000	3.6351	1.6521	0.22576	50	1.64	008.08	0
008	Fiber (%)	0407	6.8629	0.05410	3.6351	1.6521	0.22576	50	1.95	008.08	0
008	Fiber (%)	2012	7.1050	0.35000	3.6351	1.6521	0.22576	50	2.10	008.08	0
008	Fiber (%)	0848	8.6550	1.8300	3.6351	1.6521	0.22576	50	3.04	008.08	1
009	Fiber (%)	2004	6.7100	0.02000	11.455	4.5362	0.72567	48	-1.05	009.09	0
009	Fiber (%)	0297	6.7300	1.0200	11.455	4.5362	0.72567	48	-1.04	009.09	0
009	Fiber (%)	0948	6.8500	0.26000	11.455	4.5362	0.72567	48	-1.02	009.09	0
009	Fiber (%)	0263	7.1050	0.05500	11.455	4.5362	0.72567	48	-0.96	009.09	0
009	Fiber (%)	0870	7.3786	0.34740	11.455	4.5362	0.72567	48	-0.90	009.09	0
009	Fiber (%)	0353	7.5650	0.53000	11.455	4.5362	0.72567	48	-0.86	009.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
009	Fiber (%)	0148	7.6750	0.81000	11.455	4.5362	0.72567	48	-0.83	009.09	0
009	Fiber (%)	0015	7.8100	0.76000	11.455	4.5362	0.72567	48	-0.80	009.09	0
009	Fiber (%)	0581	7.9150	0.25000	11.455	4.5362	0.72567	48	-0.78	009.09	0
009	Fiber (%)	0619	8.1000	0.28000	11.455	4.5362	0.72567	48	-0.74	009.09	0
009	Fiber (%)	0226	8.2500	0.10000	11.455	4.5362	0.72567	48	-0.71	009.07	0
009	Fiber (%)	0939	8.2600	0.66000	11.455	4.5362	0.72567	48	-0.70	009.09	0
009	Fiber (%)	0861	8.2800	3.6600	11.455	4.5362	0.72567	48	-0.70	009.09	0
009	Fiber (%)	0693	8.4445	0.06700	11.455	4.5362	0.72567	48	-0.66	009.09	0
009	Fiber (%)	0638	8.5000	0.20000	11.455	4.5362	0.72567	48	-0.65	009.09	0
009	Fiber (%)	0956	8.7000	1.2000	11.455	4.5362	0.72567	48	-0.61	009.09	0
009	Fiber (%)	0646	8.7400	0.58000	11.455	4.5362	0.72567	48	-0.60	009.09	0
009	Fiber (%)	0918	8.8462	0.26580	11.455	4.5362	0.72567	48	-0.58	009.09	0
009	Fiber (%)	0096	8.8900	1.1000	11.455	4.5362	0.72567	48	-0.57	009.09	0
009	Fiber (%)	0728	8.9400	0.98000	11.455	4.5362	0.72567	48	-0.55	009.99	0
009	Fiber (%)	0001	9.2500	0.84000	11.455	4.5362	0.72567	48	-0.49	009.09	0
009	Fiber (%)	0675	9.4150	0.75000	11.455	4.5362	0.72567	48	-0.45	009.07	0
009	Fiber (%)	0309	9.4534	0.29940	11.455	4.5362	0.72567	48	-0.44	009.07	0
009	Fiber (%)	0536	9.5550	0.73000	11.455	4.5362	0.72567	48	-0.42	009.09	0
009	Fiber (%)	0066	10.035	0.27000	11.455	4.5362	0.72567	48	-0.31	009.09	0
009	Fiber (%)	0083	10.190	0.56000	11.455	4.5362	0.72567	48	-0.28	009.09	0
009	Fiber (%)	0357	10.750	0.10000	11.455	4.5362	0.72567	48	-0.16	009.09	0
009	Fiber (%)	0358	11.140	0.42000	11.455	4.5362	0.72567	48	-0.07	009.09	0
009	Fiber (%)	2129	11.370	0.36000	11.455	4.5362	0.72567	48	-0.02	009.09	0
009	Fiber (%)	0265	11.400	0.60000	11.455	4.5362	0.72567	48	-0.01	009.09	0
009	Fiber (%)	0610	12.100	0.80000	11.455	4.5362	0.72567	48	0.14	009.07	0
009	Fiber (%)	0294	12.350	0.30000	11.455	4.5362	0.72567	48	0.20	009.09	0
009	Fiber (%)	0049	12.355	0.27000	11.455	4.5362	0.72567	48	0.20	009.09	0
009	Fiber (%)	0598	13.010	3.3000	11.455	4.5362	0.72567	48	0.34	009.09	0
009	Fiber (%)	0354	13.130	0.10000	11.455	4.5362	0.72567	48	0.37	009.09	0
009	Fiber (%)	0037	13.900	1.2200	11.455	4.5362	0.72567	48	0.54	009.09	0
009	Fiber (%)	0510	14.500	0.20000	11.455	4.5362	0.72567	48	0.67	009.09	0
009	Fiber (%)	0278	15.400	1.0000	11.455	4.5362	0.72567	48	0.87	009.09	0
009	Fiber (%)	0590	16.495	2.1900	11.455	4.5362	0.72567	48	1.11	009.07	0
009	Fiber (%)	0407	18.316	0.06750	11.455	4.5362	0.72567	48	1.51	009.09	0
009	Fiber (%)	2012	19.490	0.36000	11.455	4.5362	0.72567	48	1.77	009.09	0
009	Fiber (%)	0413	19.900	0.00000	11.455	4.5362	0.72567	48	1.86	009.09	0
009	Fiber (%)	0676	20.400	1.3600	11.455	4.5362	0.72567	48	1.97	009.07	0
009	Fiber (%)	0726	22.820	2.0800	11.455	4.5362	0.72567	48	2.51	009.09	0
009	Fiber (%)	0208	23.050	0.10000	11.455	4.5362	0.72567	48	2.56	009.07	0
009	Fiber (%)	0190	23.850	0.10000	11.455	4.5362	0.72567	48	2.73	009.09	0
009	Fiber (%)	0045	24.600	0.20000	11.455	4.5362	0.72567	48	2.90	009.09	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
009	Fiber (%)	0848	33.925	3.1100	11.455	4.5362	0.72567	48	4.95	009.09	0
009	Fiber (%)	0910	11.450	5.1000	11.455	4.5362	0.72567	48	0.00	009.07	1
009	Fiber (%)	0720	0.00000	0.00000	11.455	4.5362	0.72567	48		009.99	4
010	Moisture (%)	0852	3.3800	0.02000	4.9070	0.43517	0.05605	26	-3.51	010.11	0
010	Moisture (%)	2105	3.7350	0.05000	4.9070	0.43517	0.05605	26	-2.69	010.11	0
010	Moisture (%)	0546	4.1850	0.03000	4.9070	0.43517	0.05605	26	-1.66	010.99	0
010	Moisture (%)	0843	4.2600	0.18000	4.9070	0.43517	0.05605	26	-1.49	010.03	0
010	Moisture (%)	0028	4.4650	0.01000	4.9070	0.43517	0.05605	26	-1.02	010.99	0
010	Moisture (%)	0722	4.6607	0.07720	4.9070	0.43517	0.05605	26	-0.57	010.99	0
010	Moisture (%)	2004	4.7300	0.04000	4.9070	0.43517	0.05605	26	-0.41	010.99	0
010	Moisture (%)	0884	4.8000	0.00000	4.9070	0.43517	0.05605	26	-0.25	010.99	0
010	Moisture (%)	0921	4.8200	0.00000	4.9070	0.43517	0.05605	26	-0.20	010.99	0
010	Moisture (%)	2145	4.8800	0.00000	4.9070	0.43517	0.05605	26	-0.06	010.99	0
010	Moisture (%)	2089	4.8900	0.04000	4.9070	0.43517	0.05605	26	-0.04	010.99	0
010	Moisture (%)	0939	4.9100	0.16000	4.9070	0.43517	0.05605	26	0.01	010.99	0
010	Moisture (%)	0208	4.9300	0.06000	4.9070	0.43517	0.05605	26	0.05	010.03	0
010	Moisture (%)	0652	4.9500	0.10000	4.9070	0.43517	0.05605	26	0.10	010.99	0
010	Moisture (%)	0692	4.9500	0.10000	4.9070	0.43517	0.05605	26	0.10	010.99	0
010	Moisture (%)	0948	4.9550	0.03000	4.9070	0.43517	0.05605	26	0.11	010.99	0
010	Moisture (%)	2129	4.9700	0.04000	4.9070	0.43517	0.05605	26	0.14	010.99	0
010	Moisture (%)	2150	5.0600	0.04000	4.9070	0.43517	0.05605	26	0.35	010.99	0
010	Moisture (%)	0405	5.0700	0.14000	4.9070	0.43517	0.05605	26	0.37	010.99	0
010	Moisture (%)	2144	5.0700	0.02000	4.9070	0.43517	0.05605	26	0.37	010.99	0
010	Moisture (%)	0720	5.2500	0.06000	4.9070	0.43517	0.05605	26	0.79	010.11	0
010	Moisture (%)	0726	5.3050	0.01000	4.9070	0.43517	0.05605	26	0.91	010.99	0
010	Moisture (%)	0190	5.3750	0.01000	4.9070	0.43517	0.05605	26	1.08	010.99	0
010	Moisture (%)	0152	5.4000	0.20000	4.9070	0.43517	0.05605	26	1.13	010.11	0
010	Moisture (%)	0889	5.7800	0.02000	4.9070	0.43517	0.05605	26	2.01	010.11	0
010	Moisture (%)	2076	5.8700	0.02000	4.9070	0.43517	0.05605	26	2.21	010.99	0
011	Loss on Drying (%)	0345	4.1600	0.08000	5.3722	0.29347	0.10992	76	-4.13	011.01	0
011	Loss on Drying (%)	0536	4.5050	0.05000	5.3722	0.29347	0.10992	76	-2.96	011.01	0
011	Loss on Drying (%)	0848	4.6450	0.07000	5.3722	0.29347	0.10992	76	-2.48	011.01	0
011	Loss on Drying (%)	0510	4.9000	0.20000	5.3722	0.29347	0.10992	76	-1.61	011.01	0
011	Loss on Drying (%)	2073	4.9350	0.01000	5.3722	0.29347	0.10992	76	-1.49	011.01	0
011	Loss on Drying (%)	0108	4.9750	0.05000	5.3722	0.29347	0.10992	76	-1.35	011.01	0
011	Loss on Drying (%)	0574	4.9850	0.05000	5.3722	0.29347	0.10992	76	-1.32	011.01	0
011	Loss on Drying (%)	0874	5.0100	0.04000	5.3722	0.29347	0.10992	76	-1.23	011.01	0
011	Loss on Drying (%)	0298	5.0200	0.02000	5.3722	0.29347	0.10992	76	-1.20	011.01	0
011	Loss on Drying (%)	0953	5.0350	0.05000	5.3722	0.29347	0.10992	76	-1.15	011.01	0
011	Loss on Drying (%)	0857	5.0500	0.10000	5.3722	0.29347	0.10992	76	-1.10	011.99	0
011	Loss on Drying (%)	0511	5.1200	0.52000	5.3722	0.29347	0.10992	76	-0.86	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
011	Loss on Drying (%)	0942	5.1400	0.06000	5.3722	0.29347	0.10992	76	-0.79	011.02	0
011	Loss on Drying (%)	0553	5.1550	0.13000	5.3722	0.29347	0.10992	76	-0.74	011.01	0
011	Loss on Drying (%)	0598	5.1750	0.03000	5.3722	0.29347	0.10992	76	-0.67	011.01	0
011	Loss on Drying (%)	0208	5.1800	0.12000	5.3722	0.29347	0.10992	76	-0.66	011.01	0
011	Loss on Drying (%)	0682	5.1900	0.00000	5.3722	0.29347	0.10992	76	-0.62	011.01	0
011	Loss on Drying (%)	0226	5.2000	0.00000	5.3722	0.29347	0.10992	76	-0.59	011.01	0
011	Loss on Drying (%)	0100	5.2050	0.11000	5.3722	0.29347	0.10992	76	-0.57	011.01	0
011	Loss on Drying (%)	0870	5.2100	0.10890	5.3722	0.29347	0.10992	76	-0.55	011.01	0
011	Loss on Drying (%)	0033	5.2100	0.04000	5.3722	0.29347	0.10992	76	-0.55	011.01	0
011	Loss on Drying (%)	0563	5.2118	0.41020	5.3722	0.29347	0.10992	76	-0.55	011.01	0
011	Loss on Drying (%)	0354	5.2150	0.01000	5.3722	0.29347	0.10992	76	-0.54	011.01	0
011	Loss on Drying (%)	0541	5.2300	0.88000	5.3722	0.29347	0.10992	76	-0.48	011.01	0
011	Loss on Drying (%)	0675	5.2300	0.02000	5.3722	0.29347	0.10992	76	-0.48	011.01	0
011	Loss on Drying (%)	0615	5.2800	0.06000	5.3722	0.29347	0.10992	76	-0.31	011.01	0
011	Loss on Drying (%)	2006	5.2850	0.03000	5.3722	0.29347	0.10992	76	-0.30	011.01	0
011	Loss on Drying (%)	0425	5.2850	0.05000	5.3722	0.29347	0.10992	76	-0.30	011.02	0
011	Loss on Drying (%)	0202	5.2900	0.08000	5.3722	0.29347	0.10992	76	-0.28	011.01	0
011	Loss on Drying (%)	2031	5.2950	0.83890	5.3722	0.29347	0.10992	76	-0.26	011.01	0
011	Loss on Drying (%)	0229	5.2950	0.01000	5.3722	0.29347	0.10992	76	-0.26	011.01	0
011	Loss on Drying (%)	0876	5.2950	0.01000	5.3722	0.29347	0.10992	76	-0.26	011.01	0
011	Loss on Drying (%)	0294	5.3000	0.00000	5.3722	0.29347	0.10992	76	-0.25	011.01	0
011	Loss on Drying (%)	0233	5.3100	0.02000	5.3722	0.29347	0.10992	76	-0.21	011.01	0
011	Loss on Drying (%)	0589	5.3100	0.60000	5.3722	0.29347	0.10992	76	-0.21	011.01	0
011	Loss on Drying (%)	0650	5.3100	0.24000	5.3722	0.29347	0.10992	76	-0.21	011.01	0
011	Loss on Drying (%)	0407	5.3102	0.00170	5.3722	0.29347	0.10992	76	-0.21	011.01	0
011	Loss on Drying (%)	0843	5.3200	0.02000	5.3722	0.29347	0.10992	76	-0.18	011.02	0
011	Loss on Drying (%)	0148	5.3250	0.01000	5.3722	0.29347	0.10992	76	-0.16	011.01	0
011	Loss on Drying (%)	0171	5.3300	0.12000	5.3722	0.29347	0.10992	76	-0.14	011.01	0
011	Loss on Drying (%)	0653	5.3300	0.00000	5.3722	0.29347	0.10992	76	-0.14	011.01	0
011	Loss on Drying (%)	0674	5.3300	0.00000	5.3722	0.29347	0.10992	76	-0.14	011.01	0
011	Loss on Drying (%)	2033	5.3450	0.03000	5.3722	0.29347	0.10992	76	-0.09	011.01	0
011	Loss on Drying (%)	0144	5.3600	0.08000	5.3722	0.29347	0.10992	76	-0.04	011.01	0
011	Loss on Drying (%)	0722	5.3819	0.21490	5.3722	0.29347	0.10992	76	0.03	011.01	0
011	Loss on Drying (%)	0968	5.3915	0.02500	5.3722	0.29347	0.10992	76	0.07	011.01	0
011	Loss on Drying (%)	0242	5.4050	0.01000	5.3722	0.29347	0.10992	76	0.11	011.01	0
011	Loss on Drying (%)	0552	5.4200	0.32000	5.3722	0.29347	0.10992	76	0.16	011.01	0
011	Loss on Drying (%)	0520	5.4500	0.10000	5.3722	0.29347	0.10992	76	0.27	011.01	0
011	Loss on Drying (%)	2097	5.4560	0.07150	5.3722	0.29347	0.10992	76	0.29	011.01	0
011	Loss on Drying (%)	0098	5.4650	0.07000	5.3722	0.29347	0.10992	76	0.32	011.01	0
011	Loss on Drying (%)	0263	5.4650	0.07000	5.3722	0.29347	0.10992	76	0.32	011.01	0
011	Loss on Drying (%)	0646	5.5050	0.03000	5.3722	0.29347	0.10992	76	0.45	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
011	Loss on Drying (%)	0066	5.5200	0.12000	5.3722	0.29347	0.10992	76	0.50	011.01	0
011	Loss on Drying (%)	0051	5.5250	0.21000	5.3722	0.29347	0.10992	76	0.52	011.01	0
011	Loss on Drying (%)	0622	5.5690	0.04080	5.3722	0.29347	0.10992	76	0.67	011.01	0
011	Loss on Drying (%)	0205	5.6000	0.16000	5.3722	0.29347	0.10992	76	0.78	011.01	0
011	Loss on Drying (%)	0539	5.6000	0.22000	5.3722	0.29347	0.10992	76	0.78	011.01	0
011	Loss on Drying (%)	0123	5.6050	0.03000	5.3722	0.29347	0.10992	76	0.79	011.01	0
011	Loss on Drying (%)	0728	5.6155	0.01100	5.3722	0.29347	0.10992	76	0.83	011.01	0
011	Loss on Drying (%)	0164	5.6300	0.00000	5.3722	0.29347	0.10992	76	0.88	011.01	0
011	Loss on Drying (%)	0623	5.6340	0.30800	5.3722	0.29347	0.10992	76	0.89	011.01	0
011	Loss on Drying (%)	0047	5.6500	0.10000	5.3722	0.29347	0.10992	76	0.95	011.01	0
011	Loss on Drying (%)	0309	5.6600	0.10000	5.3722	0.29347	0.10992	76	0.98	011.01	0
011	Loss on Drying (%)	0955	5.6750	0.17000	5.3722	0.29347	0.10992	76	1.03	011.01	0
011	Loss on Drying (%)	0859	5.7195	0.04100	5.3722	0.29347	0.10992	76	1.18	011.01	0
011	Loss on Drying (%)	0685	5.7500	0.02000	5.3722	0.29347	0.10992	76	1.29	011.01	0
011	Loss on Drying (%)	2022	5.7950	0.03000	5.3722	0.29347	0.10992	76	1.44	011.01	0
011	Loss on Drying (%)	0265	5.8000	0.00000	5.3722	0.29347	0.10992	76	1.46	011.01	0
011	Loss on Drying (%)	2068	5.8200	0.00000	5.3722	0.29347	0.10992	76	1.53	011.01	0
011	Loss on Drying (%)	0004	5.8250	0.15000	5.3722	0.29347	0.10992	76	1.54	011.01	0
011	Loss on Drying (%)	0358	5.8550	0.05000	5.3722	0.29347	0.10992	76	1.65	011.01	0
011	Loss on Drying (%)	0861	5.9000	0.20000	5.3722	0.29347	0.10992	76	1.80	011.01	0
011	Loss on Drying (%)	0175	5.9500	0.10000	5.3722	0.29347	0.10992	76	1.97	011.01	0
011	Loss on Drying (%)	0651	5.9960	0.00200	5.3722	0.29347	0.10992	76	2.13	011.01	0
011	Loss on Drying (%)	0852	6.0300	0.02000	5.3722	0.29347	0.10992	76	2.24	011.99	0
011	Loss on Drying (%)	0880	5.1000	3.8400	5.3722	0.29347	0.10992	76	-0.93	011.01	1
012	Starch (%)	0918	5.0250	0.22200	7.9161	1.2901	0.21372	26	-2.24	012.01	0
012	Starch (%)	0939	5.4900	0.66000	7.9161	1.2901	0.21372	26	-1.88	012.00	0
012	Starch (%)	0638	5.5000	0.00000	7.9161	1.2901	0.21372	26	-1.87	012.00	0
012	Starch (%)	2006	6.4700	0.04000	7.9161	1.2901	0.21372	26	-1.12	012.00	0
012	Starch (%)	0910	6.7000	0.60000	7.9161	1.2901	0.21372	26	-0.94	012.00	0
012	Starch (%)	0552	6.7400	0.22000	7.9161	1.2901	0.21372	26	-0.91	012.00	0
012	Starch (%)	0885	6.7420	0.22760	7.9161	1.2901	0.21372	26	-0.91	012.03	0
012	Starch (%)	0164	7.1000	0.20000	7.9161	1.2901	0.21372	26	-0.63	012.00	0
012	Starch (%)	2009	7.3837	0.08370	7.9161	1.2901	0.21372	26	-0.41	012.00	0
012	Starch (%)	0510	7.8500	0.50000	7.9161	1.2901	0.21372	26	-0.05	012.04	0
012	Starch (%)	0045	7.9800	0.16000	7.9161	1.2901	0.21372	26	0.05	012.01	0
012	Starch (%)	0407	8.0800	0.00000	7.9161	1.2901	0.21372	26	0.13	012.03	0
012	Starch (%)	0676	8.1750	0.07000	7.9161	1.2901	0.21372	26	0.20	012.01	0
012	Starch (%)	0870	8.3188	0.10340	7.9161	1.2901	0.21372	26	0.31	012.01	0
012	Starch (%)	0354	8.3600	0.20000	7.9161	1.2901	0.21372	26	0.34	012.00	0
012	Starch (%)	2004	8.3950	0.17000	7.9161	1.2901	0.21372	26	0.37	012.01	0
012	Starch (%)	0297	8.4300	0.10000	7.9161	1.2901	0.21372	26	0.40	012.03	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
012	Starch (%)	0693	8.4500	0.30000	7.9161	1.2901	0.21372	26	0.41	012.01	0
012	Starch (%)	0353	8.6500	0.02000	7.9161	1.2901	0.21372	26	0.57	012.04	0
012	Starch (%)	0956	8.7500	0.30000	7.9161	1.2901	0.21372	26	0.65	012.01	0
012	Starch (%)	0164	8.8000	0.40000	7.9161	1.2901	0.21372	26	0.69	012.01	0
012	Starch (%)	0278	8.8000	0.20000	7.9161	1.2901	0.21372	26	0.69	012.04	0
012	Starch (%)	0208	8.8550	0.21000	7.9161	1.2901	0.21372	26	0.73	012.04	0
012	Starch (%)	0265	9.2000	0.00000	7.9161	1.2901	0.21372	26	1.00	012.01	0
012	Starch (%)	0948	9.7950	0.37000	7.9161	1.2901	0.21372	26	1.46	012.04	0
012	Starch (%)	0610	18.700	0.20000	7.9161	1.2901	0.21372	26	8.36	012.00	0
012	Starch (%)	0720	0.00000	0.00000	7.9161	1.2901	0.21372	26		012.11	4
013	Fat (%)	0407	17.015	0.30060	19.685	0.89169	0.27790	62	-2.99	013.13	0
013	Fat (%)	0036	17.219	0.03730	19.685	0.89169	0.27790	62	-2.77	013.10	0
013	Fat (%)	0047	17.955	0.05000	19.685	0.89169	0.27790	62	-1.94	013.08	0
013	Fat (%)	0629	17.985	0.37000	19.685	0.89169	0.27790	62	-1.91	013.10	0
013	Fat (%)	0615	18.550	0.70000	19.685	0.89169	0.27790	62	-1.27	013.00	0
013	Fat (%)	0242	18.560	0.42000	19.685	0.89169	0.27790	62	-1.26	013.00	0
013	Fat (%)	0297	18.590	0.36000	19.685	0.89169	0.27790	62	-1.23	013.00	0
013	Fat (%)	0843	18.665	0.53000	19.685	0.89169	0.27790	62	-1.14	013.13	0
013	Fat (%)	0353	18.700	0.44000	19.685	0.89169	0.27790	62	-1.10	013.10	0
013	Fat (%)	0968	18.797	0.09400	19.685	0.89169	0.27790	62	-1.00	013.13	0
013	Fat (%)	0643	18.880	0.14000	19.685	0.89169	0.27790	62	-0.90	013.02	0
013	Fat (%)	0505	18.885	0.07000	19.685	0.89169	0.27790	62	-0.90	013.02	0
013	Fat (%)	2006	18.975	0.01000	19.685	0.89169	0.27790	62	-0.80	013.10	0
013	Fat (%)	0948	19.010	0.04000	19.685	0.89169	0.27790	62	-0.76	013.00	0
013	Fat (%)	0618	19.027	0.37260	19.685	0.89169	0.27790	62	-0.74	013.08	0
013	Fat (%)	0541	19.035	0.09000	19.685	0.89169	0.27790	62	-0.73	013.00	0
013	Fat (%)	0100	19.055	0.01000	19.685	0.89169	0.27790	62	-0.71	013.02	0
013	Fat (%)	0955	19.100	0.20000	19.685	0.89169	0.27790	62	-0.66	013.13	0
013	Fat (%)	0011	19.215	0.49000	19.685	0.89169	0.27790	62	-0.53	013.02	0
013	Fat (%)	0083	19.295	0.21000	19.685	0.89169	0.27790	62	-0.44	013.00	0
013	Fat (%)	0511	19.345	0.25000	19.685	0.89169	0.27790	62	-0.38	013.02	0
013	Fat (%)	0265	19.350	0.50000	19.685	0.89169	0.27790	62	-0.38	013.13	0
013	Fat (%)	0939	19.425	0.11000	19.685	0.89169	0.27790	62	-0.29	013.00	0
013	Fat (%)	0035	19.465	1.2700	19.685	0.89169	0.27790	62	-0.25	013.00	0
013	Fat (%)	0098	19.505	0.11000	19.685	0.89169	0.27790	62	-0.20	013.00	0
013	Fat (%)	0610	19.550	0.44000	19.685	0.89169	0.27790	62	-0.15	013.10	0
013	Fat (%)	0208	19.600	0.80000	19.685	0.89169	0.27790	62	-0.10	013.02	0
013	Fat (%)	0139	19.615	0.03000	19.685	0.89169	0.27790	62	-0.08	013.00	0
013	Fat (%)	0884	19.650	0.30000	19.685	0.89169	0.27790	62	-0.04	013.00	0
013	Fat (%)	2004	19.650	0.70000	19.685	0.89169	0.27790	62	-0.04	013.00	0
013	Fat (%)	2149	19.675	0.45000	19.685	0.89169	0.27790	62	-0.01	013.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
013	Fat (%)	0652	19.700	0.00000	19.685	0.89169	0.27790	62	0.02	013.00	0
013	Fat (%)	2076	19.702	0.40310	19.685	0.89169	0.27790	62	0.02	013.00	0
013	Fat (%)	0870	19.706	0.60490	19.685	0.89169	0.27790	62	0.02	013.02	0
013	Fat (%)	0638	19.750	0.10000	19.685	0.89169	0.27790	62	0.07	013.10	0
013	Fat (%)	0019	19.840	0.06000	19.685	0.89169	0.27790	62	0.17	013.13	0
013	Fat (%)	2073	19.865	0.09000	19.685	0.89169	0.27790	62	0.20	013.00	0
013	Fat (%)	0229	19.895	0.53000	19.685	0.89169	0.27790	62	0.24	013.00	0
013	Fat (%)	0164	19.910	0.06000	19.685	0.89169	0.27790	62	0.25	013.02	0
013	Fat (%)	0017	19.950	0.70000	19.685	0.89169	0.27790	62	0.30	013.02	0
013	Fat (%)	0650	19.960	0.12000	19.685	0.89169	0.27790	62	0.31	013.02	0
013	Fat (%)	0553	19.970	0.14000	19.685	0.89169	0.27790	62	0.32	013.02	0
013	Fat (%)	0019	20.085	0.27000	19.685	0.89169	0.27790	62	0.45	013.02	0
013	Fat (%)	0510	20.100	0.20000	19.685	0.89169	0.27790	62	0.47	013.02	0
013	Fat (%)	0309	20.108	0.17160	19.685	0.89169	0.27790	62	0.47	013.00	0
013	Fat (%)	0034	20.175	0.23000	19.685	0.89169	0.27790	62	0.55	013.00	0
013	Fat (%)	0010	20.200	0.20000	19.685	0.89169	0.27790	62	0.58	013.02	0
013	Fat (%)	2053	20.220	0.22000	19.685	0.89169	0.27790	62	0.60	013.00	0
013	Fat (%)	0171	20.225	0.11000	19.685	0.89169	0.27790	62	0.61	013.02	0
013	Fat (%)	2150	20.235	0.27000	19.685	0.89169	0.27790	62	0.62	013.02	0
013	Fat (%)	0148	20.445	0.03000	19.685	0.89169	0.27790	62	0.85	013.02	0
013	Fat (%)	0202	20.470	0.26000	19.685	0.89169	0.27790	62	0.88	013.00	0
013	Fat (%)	0027	20.665	0.23000	19.685	0.89169	0.27790	62	1.10	013.13	0
013	Fat (%)	0354	20.820	0.46000	19.685	0.89169	0.27790	62	1.27	013.02	0
013	Fat (%)	0682	20.840	0.00000	19.685	0.89169	0.27790	62	1.30	013.02	0
013	Fat (%)	0675	20.955	0.19000	19.685	0.89169	0.27790	62	1.42	013.02	0
013	Fat (%)	0618	21.015	0.34600	19.685	0.89169	0.27790	62	1.49	013.00	0
013	Fat (%)	0861	21.120	0.16000	19.685	0.89169	0.27790	62	1.61	013.02	0
013	Fat (%)	0581	21.575	0.25000	19.685	0.89169	0.27790	62	2.12	013.13	0
013	Fat (%)	0676	21.690	0.14000	19.685	0.89169	0.27790	62	2.25	013.02	0
013	Fat (%)	0689	22.150	0.50000	19.685	0.89169	0.27790	62	2.76	013.00	0
013	Fat (%)	0003	22.215	0.29000	19.685	0.89169	0.27790	62	2.84	013.13	0
013	Fat (%)	0874	16.900	1.7200	19.685	0.89169	0.27790	62	-3.12	013.02	1
013	Fat (%)	0910	19.170	1.5600	19.685	0.89169	0.27790	62	-0.58	013.00	1
015	Aluminum (mg / kg (ppm))	0049	213.06	2.6200	349.78	42.227	13.787	15	-3.24	015.41	0
015	Aluminum (mg / kg (ppm))	0186	245.00	20.000	349.78	42.227	13.787	15	-2.48	015.52	0
015	Aluminum (mg / kg (ppm))	0918	321.85	25.586	349.78	42.227	13.787	15	-0.66	015.43	0
015	Aluminum (mg / kg (ppm))	0164	330.50	11.000	349.78	42.227	13.787	15	-0.46	015.41	0
015	Aluminum (mg / kg (ppm))	0345	334.50	5.0000	349.78	42.227	13.787	15	-0.36	015.43	0
015	Aluminum (mg / kg (ppm))	0510	339.00	2.0000	349.78	42.227	13.787	15	-0.26	015.43	0
015	Aluminum (mg / kg (ppm))	0171	344.10	2.8000	349.78	42.227	13.787	15	-0.13	015.41	0
015	Aluminum (mg / kg (ppm))	0520	352.50	11.000	349.78	42.227	13.787	15	0.06	015.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
015	Aluminum (mg / kg (ppm))	0169	354.50	9.0000	349.78	42.227	13.787	15	0.11	015.43	0
015	Aluminum (mg / kg (ppm))	0353	354.50	1.0000	349.78	42.227	13.787	15	0.11	015.43	0
015	Aluminum (mg / kg (ppm))	0560	361.10	18.800	349.78	42.227	13.787	15	0.27	015.42	0
015	Aluminum (mg / kg (ppm))	0413	366.00	34.000	349.78	42.227	13.787	15	0.38	015.53	0
015	Aluminum (mg / kg (ppm))	0297	389.00	16.000	349.78	42.227	13.787	15	0.93	015.43	0
015	Aluminum (mg / kg (ppm))	0553	415.50	1.0000	349.78	42.227	13.787	15	1.56	015.53	0
015	Aluminum (mg / kg (ppm))	2033	419.00	47.000	349.78	42.227	13.787	15	1.64	015.43	0
017	Boron (mg / kg (ppm))	0229	1.6650	0.95000	2.6262	0.78238	0.30211	14	-1.23	017.41	0
017	Boron (mg / kg (ppm))	0226	1.8250	0.03000	2.6262	0.78238	0.30211	14	-1.02	017.41	0
017	Boron (mg / kg (ppm))	0045	2.0100	0.22000	2.6262	0.78238	0.30211	14	-0.79	017.42	0
017	Boron (mg / kg (ppm))	0560	2.0400	0.36000	2.6262	0.78238	0.30211	14	-0.75	017.42	0
017	Boron (mg / kg (ppm))	0358	2.3000	0.14000	2.6262	0.78238	0.30211	14	-0.42	017.41	0
017	Boron (mg / kg (ppm))	0693	2.3245	0.39100	2.6262	0.78238	0.30211	14	-0.39	017.42	0
017	Boron (mg / kg (ppm))	0345	2.3500	0.30000	2.6262	0.78238	0.30211	14	-0.35	017.43	0
017	Boron (mg / kg (ppm))	0413	2.5500	0.30000	2.6262	0.78238	0.30211	14	-0.10	017.53	0
017	Boron (mg / kg (ppm))	0049	2.6450	0.25000	2.6262	0.78238	0.30211	14	0.02	017.41	0
017	Boron (mg / kg (ppm))	2033	3.0000	0.00000	2.6262	0.78238	0.30211	14	0.48	017.43	0
017	Boron (mg / kg (ppm))	0918	3.0569	0.00860	2.6262	0.78238	0.30211	14	0.55	017.43	0
017	Boron (mg / kg (ppm))	0353	3.5000	1.0000	2.6262	0.78238	0.30211	14	1.12	017.43	0
017	Boron (mg / kg (ppm))	0510	3.7000	0.00000	2.6262	0.78238	0.30211	14	1.37	017.43	0
017	Boron (mg / kg (ppm))	0294	6.2600	0.28000	2.6262	0.78238	0.30211	14	4.64	017.42	0
017	Boron (mg / kg (ppm))	0297	0.00000	0.00000	2.6262	0.78238	0.30211	14		017.43	4
019	Calcium (%)	0889	1.9005	0.07350	2.7987	0.15806	0.06679	118	-5.68	019.99	0
019	Calcium (%)	0536	2.3300	0.26000	2.7987	0.15806	0.06679	118	-2.97	019.02	0
019	Calcium (%)	0964	2.3561	0.31460	2.7987	0.15806	0.06679	118	-2.80	019.43	0
019	Calcium (%)	0520	2.4400	0.32000	2.7987	0.15806	0.06679	118	-2.27	019.41	0
019	Calcium (%)	0629	2.4550	0.01000	2.7987	0.15806	0.06679	118	-2.17	019.08	0
019	Calcium (%)	0676	2.4950	0.07000	2.7987	0.15806	0.06679	118	-1.92	019.99	0
019	Calcium (%)	0233	2.5000	0.00000	2.7987	0.15806	0.06679	118	-1.89	019.31	0
019	Calcium (%)	0948	2.5115	0.00300	2.7987	0.15806	0.06679	118	-1.82	019.31	0
019	Calcium (%)	0148	2.5450	0.07000	2.7987	0.15806	0.06679	118	-1.61	019.41	0
019	Calcium (%)	0144	2.5550	0.09000	2.7987	0.15806	0.06679	118	-1.54	019.41	0
019	Calcium (%)	0051	2.5800	0.06000	2.7987	0.15806	0.06679	118	-1.38	019.41	0
019	Calcium (%)	0154	2.5823	0.03470	2.7987	0.15806	0.06679	118	-1.37	019.52	0
019	Calcium (%)	0622	2.6039	0.03110	2.7987	0.15806	0.06679	118	-1.23	019.00	0
019	Calcium (%)	0623	2.6150	0.10200	2.7987	0.15806	0.06679	118	-1.16	019.00	0
019	Calcium (%)	2006	2.6225	0.03300	2.7987	0.15806	0.06679	118	-1.11	019.00	0
019	Calcium (%)	0674	2.6300	0.00000	2.7987	0.15806	0.06679	118	-1.07	019.31	0
019	Calcium (%)	0045	2.6350	0.05000	2.7987	0.15806	0.06679	118	-1.04	019.42	0
019	Calcium (%)	0940	2.6400	0.02000	2.7987	0.15806	0.06679	118	-1.00	019.08	0
019	Calcium (%)	0089	2.6450	0.01000	2.7987	0.15806	0.06679	118	-0.97	019.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	2022	2.6500	0.02000	2.7987	0.15806	0.06679	118	-0.94	019.31	0
019	Calcium (%)	0175	2.6550	0.03000	2.7987	0.15806	0.06679	118	-0.91	019.31	0
019	Calcium (%)	0042	2.6600	0.26000	2.7987	0.15806	0.06679	118	-0.88	019.42	0
019	Calcium (%)	0263	2.6638	0.03200	2.7987	0.15806	0.06679	118	-0.85	019.42	0
019	Calcium (%)	0536	2.6700	0.06000	2.7987	0.15806	0.06679	118	-0.81	019.31	0
019	Calcium (%)	2004	2.6700	0.02000	2.7987	0.15806	0.06679	118	-0.81	019.44	0
019	Calcium (%)	0510	2.6850	0.01000	2.7987	0.15806	0.06679	118	-0.72	019.43	0
019	Calcium (%)	0675	2.6950	0.11000	2.7987	0.15806	0.06679	118	-0.66	019.43	0
019	Calcium (%)	0108	2.7000	0.08000	2.7987	0.15806	0.06679	118	-0.62	019.31	0
019	Calcium (%)	2006	2.7045	0.13100	2.7987	0.15806	0.06679	118	-0.60	019.09	0
019	Calcium (%)	0202	2.7065	0.02500	2.7987	0.15806	0.06679	118	-0.58	019.43	0
019	Calcium (%)	0939	2.7100	0.08000	2.7987	0.15806	0.06679	118	-0.56	019.31	0
019	Calcium (%)	0171	2.7100	0.00000	2.7987	0.15806	0.06679	118	-0.56	019.41	0
019	Calcium (%)	0123	2.7150	0.05000	2.7987	0.15806	0.06679	118	-0.53	019.41	0
019	Calcium (%)	0968	2.7150	0.01000	2.7987	0.15806	0.06679	118	-0.53	019.43	0
019	Calcium (%)	0590	2.7150	0.01000	2.7987	0.15806	0.06679	118	-0.53	019.99	0
019	Calcium (%)	0874	2.7200	0.04000	2.7987	0.15806	0.06679	118	-0.50	019.31	0
019	Calcium (%)	0139	2.7250	0.03000	2.7987	0.15806	0.06679	118	-0.47	019.41	0
019	Calcium (%)	0651	2.7260	0.03200	2.7987	0.15806	0.06679	118	-0.46	019.00	0
019	Calcium (%)	0560	2.7261	0.01140	2.7987	0.15806	0.06679	118	-0.46	019.42	0
019	Calcium (%)	0693	2.7300	0.22000	2.7987	0.15806	0.06679	118	-0.43	019.42	0
019	Calcium (%)	0405	2.7300	0.02000	2.7987	0.15806	0.06679	118	-0.43	019.41	0
019	Calcium (%)	0164	2.7350	0.07000	2.7987	0.15806	0.06679	118	-0.40	019.41	0
019	Calcium (%)	0653	2.7375	0.00700	2.7987	0.15806	0.06679	118	-0.39	019.31	0
019	Calcium (%)	0590	2.7450	0.13000	2.7987	0.15806	0.06679	118	-0.34	019.08	0
019	Calcium (%)	0345	2.7475	0.01500	2.7987	0.15806	0.06679	118	-0.32	019.43	0
019	Calcium (%)	0508	2.7484	0.04320	2.7987	0.15806	0.06679	118	-0.32	019.43	0
019	Calcium (%)	0186	2.7500	0.16000	2.7987	0.15806	0.06679	118	-0.31	019.52	0
019	Calcium (%)	0036	2.7600	0.00760	2.7987	0.15806	0.06679	118	-0.25	019.43	0
019	Calcium (%)	0884	2.7700	0.08000	2.7987	0.15806	0.06679	118	-0.18	019.31	0
019	Calcium (%)	2053	2.7700	0.04000	2.7987	0.15806	0.06679	118	-0.18	019.42	0
019	Calcium (%)	0849	2.7715	0.01700	2.7987	0.15806	0.06679	118	-0.17	019.08	0
019	Calcium (%)	0205	2.7750	0.07000	2.7987	0.15806	0.06679	118	-0.15	019.42	0
019	Calcium (%)	0885	2.7765	0.04980	2.7987	0.15806	0.06679	118	-0.14	019.08	0
019	Calcium (%)	0265	2.7800	0.10000	2.7987	0.15806	0.06679	118	-0.12	019.42	0
019	Calcium (%)	0552	2.7850	0.05000	2.7987	0.15806	0.06679	118	-0.09	019.00	0
019	Calcium (%)	0354	2.7900	0.02000	2.7987	0.15806	0.06679	118	-0.06	019.31	0
019	Calcium (%)	0720	2.7900	0.00000	2.7987	0.15806	0.06679	118	-0.06	019.41	0
019	Calcium (%)	2012	2.7900	0.04000	2.7987	0.15806	0.06679	118	-0.06	019.41	0
019	Calcium (%)	0019	2.8000	0.02000	2.7987	0.15806	0.06679	118	0.01	019.41	0
019	Calcium (%)	0242	2.8000	0.02000	2.7987	0.15806	0.06679	118	0.01	019.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	2062	2.8046	0.01110	2.7987	0.15806	0.06679	118	0.04	019.00	0
019	Calcium (%)	0903	2.8050	0.01000	2.7987	0.15806	0.06679	118	0.04	019.00	0
019	Calcium (%)	0683	2.8100	0.06000	2.7987	0.15806	0.06679	118	0.07	019.00	0
019	Calcium (%)	0689	2.8150	0.03000	2.7987	0.15806	0.06679	118	0.10	019.08	0
019	Calcium (%)	0298	2.8150	0.03000	2.7987	0.15806	0.06679	118	0.10	019.41	0
019	Calcium (%)	0208	2.8230	0.16000	2.7987	0.15806	0.06679	118	0.15	019.31	0
019	Calcium (%)	0722	2.8248	0.00940	2.7987	0.15806	0.06679	118	0.16	019.31	0
019	Calcium (%)	0918	2.8250	0.18470	2.7987	0.15806	0.06679	118	0.17	019.43	0
019	Calcium (%)	0687	2.8250	0.15000	2.7987	0.15806	0.06679	118	0.17	019.31	0
019	Calcium (%)	0861	2.8250	0.01000	2.7987	0.15806	0.06679	118	0.17	019.43	0
019	Calcium (%)	0563	2.8263	0.07010	2.7987	0.15806	0.06679	118	0.17	019.31	0
019	Calcium (%)	0297	2.8350	0.01000	2.7987	0.15806	0.06679	118	0.23	019.43	0
019	Calcium (%)	0229	2.8380	0.01000	2.7987	0.15806	0.06679	118	0.25	019.41	0
019	Calcium (%)	0413	2.8500	0.10000	2.7987	0.15806	0.06679	118	0.32	019.41	0
019	Calcium (%)	0682	2.8500	0.00000	2.7987	0.15806	0.06679	118	0.32	019.43	0
019	Calcium (%)	0895	2.8550	0.01000	2.7987	0.15806	0.06679	118	0.36	019.00	0
019	Calcium (%)	0049	2.8550	0.01000	2.7987	0.15806	0.06679	118	0.36	019.41	0
019	Calcium (%)	0186	2.8550	0.01000	2.7987	0.15806	0.06679	118	0.36	019.42	0
019	Calcium (%)	0366	2.8600	0.04000	2.7987	0.15806	0.06679	118	0.39	019.42	0
019	Calcium (%)	0553	2.8600	0.00000	2.7987	0.15806	0.06679	118	0.39	019.53	0
019	Calcium (%)	2009	2.8623	0.00930	2.7987	0.15806	0.06679	118	0.40	019.08	0
019	Calcium (%)	0142	2.8709	0.00480	2.7987	0.15806	0.06679	118	0.46	019.31	0
019	Calcium (%)	0529	2.8750	0.01000	2.7987	0.15806	0.06679	118	0.48	019.31	0
019	Calcium (%)	0294	2.8750	0.03000	2.7987	0.15806	0.06679	118	0.48	019.42	0
019	Calcium (%)	0017	2.8800	0.02000	2.7987	0.15806	0.06679	118	0.51	019.43	0
019	Calcium (%)	0034	2.8824	0.22310	2.7987	0.15806	0.06679	118	0.53	019.43	0
019	Calcium (%)	0036	2.8865	0.01680	2.7987	0.15806	0.06679	118	0.56	019.03	0
019	Calcium (%)	0037	2.8900	0.06000	2.7987	0.15806	0.06679	118	0.58	019.42	0
019	Calcium (%)	0098	2.8900	0.12000	2.7987	0.15806	0.06679	118	0.58	019.41	0
019	Calcium (%)	0033	2.8950	0.01000	2.7987	0.15806	0.06679	118	0.61	019.43	0
019	Calcium (%)	2089	2.8950	0.03000	2.7987	0.15806	0.06679	118	0.61	019.43	0
019	Calcium (%)	2033	2.9000	0.00000	2.7987	0.15806	0.06679	118	0.64	019.43	0
019	Calcium (%)	0098	2.9100	0.22000	2.7987	0.15806	0.06679	118	0.70	019.43	0
019	Calcium (%)	0357	2.9178	0.02140	2.7987	0.15806	0.06679	118	0.75	019.42	0
019	Calcium (%)	0893	2.9200	0.00000	2.7987	0.15806	0.06679	118	0.77	019.00	0
019	Calcium (%)	0897	2.9200	0.00000	2.7987	0.15806	0.06679	118	0.77	019.00	0
019	Calcium (%)	0910	2.9400	0.16000	2.7987	0.15806	0.06679	118	0.89	019.41	0
019	Calcium (%)	0278	2.9400	0.04000	2.7987	0.15806	0.06679	118	0.89	019.42	0
019	Calcium (%)	0300	2.9450	0.17000	2.7987	0.15806	0.06679	118	0.93	019.33	0
019	Calcium (%)	0083	2.9650	0.07000	2.7987	0.15806	0.06679	118	1.05	019.43	0
019	Calcium (%)	0878	2.9750	0.13000	2.7987	0.15806	0.06679	118	1.12	019.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	0572	2.9850	0.25000	2.7987	0.15806	0.06679	118	1.18	019.53	0
019	Calcium (%)	0692	2.9850	0.27000	2.7987	0.15806	0.06679	118	1.18	019.42	0
019	Calcium (%)	0358	2.9900	0.16000	2.7987	0.15806	0.06679	118	1.21	019.41	0
019	Calcium (%)	0226	3.0000	0.18000	2.7987	0.15806	0.06679	118	1.27	019.41	0
019	Calcium (%)	0006	3.0055	0.03100	2.7987	0.15806	0.06679	118	1.31	019.43	0
019	Calcium (%)	0870	3.0260	0.01600	2.7987	0.15806	0.06679	118	1.44	019.42	0
019	Calcium (%)	0100	3.0450	0.03000	2.7987	0.15806	0.06679	118	1.56	019.99	0
019	Calcium (%)	0035	3.0550	0.05000	2.7987	0.15806	0.06679	118	1.62	019.42	0
019	Calcium (%)	0035	3.0700	0.34000	2.7987	0.15806	0.06679	118	1.72	019.43	0
019	Calcium (%)	0598	3.0790	0.03400	2.7987	0.15806	0.06679	118	1.77	019.41	0
019	Calcium (%)	0425	3.1200	0.04000	2.7987	0.15806	0.06679	118	2.03	019.43	0
019	Calcium (%)	0619	3.1250	0.17000	2.7987	0.15806	0.06679	118	2.06	019.41	0
019	Calcium (%)	0413	3.1300	0.06000	2.7987	0.15806	0.06679	118	2.10	019.53	0
019	Calcium (%)	0407	3.1508	0.00650	2.7987	0.15806	0.06679	118	2.23	019.41	0
019	Calcium (%)	0353	3.2350	0.07000	2.7987	0.15806	0.06679	118	2.76	019.43	0
019	Calcium (%)	0650	3.3100	0.04000	2.7987	0.15806	0.06679	118	3.23	019.31	0
019	Calcium (%)	0852	4.0500	0.10000	2.7987	0.15806	0.06679	118	7.92	019.08	0
019	Calcium (%)	2129	2.1450	0.39000	2.7987	0.15806	0.06679	118	-4.14	019.42	1
019	Calcium (%)	0208	2.9055	0.38300	2.7987	0.15806	0.06679	118	0.68	019.41	1
019	Calcium (%)	2097	6.8951	0.03400	2.7987	0.15806	0.06679	118	25.92	019.31	2
019	Calcium (%)	0169	0.00000	0.00000	2.7987	0.15806	0.06679	118		019.32	4
021	Cobalt (mg / kg (ppm))	0345	7.2650	0.27000	9.2783	1.3558	0.70561	21	-1.48	021.43	0
021	Cobalt (mg / kg (ppm))	0169	7.6000	0.44000	9.2783	1.3558	0.70561	21	-1.24	021.43	0
021	Cobalt (mg / kg (ppm))	0164	7.8500	0.50000	9.2783	1.3558	0.70561	21	-1.05	021.31	0
021	Cobalt (mg / kg (ppm))	0693	8.1750	1.2220	9.2783	1.3558	0.70561	21	-0.81	021.42	0
021	Cobalt (mg / kg (ppm))	0043	8.2200	0.06000	9.2783	1.3558	0.70561	21	-0.78	021.43	0
021	Cobalt (mg / kg (ppm))	0171	8.2400	0.20000	9.2783	1.3558	0.70561	21	-0.77	021.41	0
021	Cobalt (mg / kg (ppm))	0689	8.2500	0.30000	9.2783	1.3558	0.70561	21	-0.76	021.31	0
021	Cobalt (mg / kg (ppm))	0508	8.6028	0.79260	9.2783	1.3558	0.70561	21	-0.50	021.43	0
021	Cobalt (mg / kg (ppm))	0045	8.7400	0.08000	9.2783	1.3558	0.70561	21	-0.40	021.42	0
021	Cobalt (mg / kg (ppm))	0560	9.2150	1.4300	9.2783	1.3558	0.70561	21	-0.05	021.42	0
021	Cobalt (mg / kg (ppm))	0939	9.4550	0.17000	9.2783	1.3558	0.70561	21	0.13	021.31	0
021	Cobalt (mg / kg (ppm))	0186	9.5000	1.0000	9.2783	1.3558	0.70561	21	0.16	021.52	0
021	Cobalt (mg / kg (ppm))	0553	9.5050	0.45000	9.2783	1.3558	0.70561	21	0.17	021.53	0
021	Cobalt (mg / kg (ppm))	0510	10.040	0.02000	9.2783	1.3558	0.70561	21	0.56	021.43	0
021	Cobalt (mg / kg (ppm))	0098	10.120	0.20000	9.2783	1.3558	0.70561	21	0.62	021.53	0
021	Cobalt (mg / kg (ppm))	0964	10.124	1.8952	9.2783	1.3558	0.70561	21	0.62	021.43	0
021	Cobalt (mg / kg (ppm))	2033	10.230	1.7400	9.2783	1.3558	0.70561	21	0.70	021.43	0
021	Cobalt (mg / kg (ppm))	2053	10.460	0.44000	9.2783	1.3558	0.70561	21	0.87	021.42	0
021	Cobalt (mg / kg (ppm))	0910	10.690	2.3800	9.2783	1.3558	0.70561	21	1.04	021.52	0
021	Cobalt (mg / kg (ppm))	0413	11.250	0.30000	9.2783	1.3558	0.70561	21	1.45	021.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
021	Cobalt (mg / kg (ppm))	0918	12.392	0.92800	9.2783	1.3558	0.70561	21	2.30	021.53	0
022	Copper (mg / kg (ppm))	0884	46.200	8.2000	75.517	4.5095	1.9532	90	-6.50	022.31	0
022	Copper (mg / kg (ppm))	0405	47.400	0.26000	75.517	4.5095	1.9532	90	-6.24	022.41	0
022	Copper (mg / kg (ppm))	0016	50.650	5.5000	75.517	4.5095	1.9532	90	-5.51	022.52	0
022	Copper (mg / kg (ppm))	0037	60.450	3.9000	75.517	4.5095	1.9532	90	-3.34	022.42	0
022	Copper (mg / kg (ppm))	0511	62.000	2.0000	75.517	4.5095	1.9532	90	-3.00	022.41	0
022	Copper (mg / kg (ppm))	0520	67.500	1.0000	75.517	4.5095	1.9532	90	-1.78	022.41	0
022	Copper (mg / kg (ppm))	0345	67.550	0.50000	75.517	4.5095	1.9532	90	-1.77	022.43	0
022	Copper (mg / kg (ppm))	0043	69.200	2.2000	75.517	4.5095	1.9532	90	-1.40	022.43	0
022	Copper (mg / kg (ppm))	0529	69.200	9.6000	75.517	4.5095	1.9532	90	-1.40	022.31	0
022	Copper (mg / kg (ppm))	0955	70.000	3.6000	75.517	4.5095	1.9532	90	-1.22	022.41	0
022	Copper (mg / kg (ppm))	0918	70.130	2.4800	75.517	4.5095	1.9532	90	-1.19	022.43	0
022	Copper (mg / kg (ppm))	0019	70.400	2.2800	75.517	4.5095	1.9532	90	-1.13	022.41	0
022	Copper (mg / kg (ppm))	0045	70.850	0.30000	75.517	4.5095	1.9532	90	-1.04	022.42	0
022	Copper (mg / kg (ppm))	0689	70.900	0.60000	75.517	4.5095	1.9532	90	-1.02	022.31	0
022	Copper (mg / kg (ppm))	0202	71.020	1.1000	75.517	4.5095	1.9532	90	-1.00	022.43	0
022	Copper (mg / kg (ppm))	0047	71.045	4.6900	75.517	4.5095	1.9532	90	-0.99	022.52	0
022	Copper (mg / kg (ppm))	2149	71.080	0.74000	75.517	4.5095	1.9532	90	-0.98	022.31	0
022	Copper (mg / kg (ppm))	0049	71.625	1.6900	75.517	4.5095	1.9532	90	-0.86	022.41	0
022	Copper (mg / kg (ppm))	0148	71.750	0.10000	75.517	4.5095	1.9532	90	-0.84	022.41	0
022	Copper (mg / kg (ppm))	0407	72.081	0.85550	75.517	4.5095	1.9532	90	-0.76	022.41	0
022	Copper (mg / kg (ppm))	0263	72.200	2.0000	75.517	4.5095	1.9532	90	-0.74	022.42	0
022	Copper (mg / kg (ppm))	0878	72.290	0.06000	75.517	4.5095	1.9532	90	-0.72	022.31	0
022	Copper (mg / kg (ppm))	0169	72.300	0.20000	75.517	4.5095	1.9532	90	-0.71	022.43	0
022	Copper (mg / kg (ppm))	0693	72.453	0.28400	75.517	4.5095	1.9532	90	-0.68	022.42	0
022	Copper (mg / kg (ppm))	0563	72.454	2.6281	75.517	4.5095	1.9532	90	-0.68	022.31	0
022	Copper (mg / kg (ppm))	0208	72.625	4.4700	75.517	4.5095	1.9532	90	-0.64	022.41	0
022	Copper (mg / kg (ppm))	0051	72.975	2.3100	75.517	4.5095	1.9532	90	-0.56	022.41	0
022	Copper (mg / kg (ppm))	0629	73.000	0.00000	75.517	4.5095	1.9532	90	-0.56	022.31	0
022	Copper (mg / kg (ppm))	0242	73.000	0.00000	75.517	4.5095	1.9532	90	-0.56	022.99	0
022	Copper (mg / kg (ppm))	2129	73.020	0.70000	75.517	4.5095	1.9532	90	-0.55	022.99	0
022	Copper (mg / kg (ppm))	0619	73.700	0.00000	75.517	4.5095	1.9532	90	-0.40	022.41	0
022	Copper (mg / kg (ppm))	0939	73.710	0.90000	75.517	4.5095	1.9532	90	-0.40	022.31	0
022	Copper (mg / kg (ppm))	0682	73.900	0.00000	75.517	4.5095	1.9532	90	-0.36	022.43	0
022	Copper (mg / kg (ppm))	0910	74.000	8.0000	75.517	4.5095	1.9532	90	-0.34	022.41	0
022	Copper (mg / kg (ppm))	2062	74.550	0.70000	75.517	4.5095	1.9532	90	-0.21	022.31	0
022	Copper (mg / kg (ppm))	0861	74.900	0.80000	75.517	4.5095	1.9532	90	-0.14	022.43	0
022	Copper (mg / kg (ppm))	0205	75.250	0.10000	75.517	4.5095	1.9532	90	-0.06	022.42	0
022	Copper (mg / kg (ppm))	0590	75.400	0.60000	75.517	4.5095	1.9532	90	-0.03	022.99	0
022	Copper (mg / kg (ppm))	0553	75.450	4.5000	75.517	4.5095	1.9532	90	-0.01	022.53	0
022	Copper (mg / kg (ppm))	0297	75.500	5.0000	75.517	4.5095	1.9532	90	0.00	022.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
022	Copper (mg / kg (ppm))	0100	75.500	1.0000	75.517	4.5095	1.9532	90	0.00	022.99	0
022	Copper (mg / kg (ppm))	0278	75.550	2.7000	75.517	4.5095	1.9532	90	0.01	022.42	0
022	Copper (mg / kg (ppm))	0598	75.600	0.42000	75.517	4.5095	1.9532	90	0.02	022.41	0
022	Copper (mg / kg (ppm))	0870	75.780	0.94000	75.517	4.5095	1.9532	90	0.06	022.42	0
022	Copper (mg / kg (ppm))	0265	76.000	0.00000	75.517	4.5095	1.9532	90	0.11	022.42	0
022	Copper (mg / kg (ppm))	0510	76.000	2.0000	75.517	4.5095	1.9532	90	0.11	022.43	0
022	Copper (mg / kg (ppm))	0720	76.225	0.67000	75.517	4.5095	1.9532	90	0.16	022.41	0
022	Copper (mg / kg (ppm))	0357	76.406	0.31500	75.517	4.5095	1.9532	90	0.20	022.42	0
022	Copper (mg / kg (ppm))	0098	76.480	6.3000	75.517	4.5095	1.9532	90	0.21	022.43	0
022	Copper (mg / kg (ppm))	0164	76.500	1.0000	75.517	4.5095	1.9532	90	0.22	022.41	0
022	Copper (mg / kg (ppm))	0035	76.500	1.0000	75.517	4.5095	1.9532	90	0.22	022.42	0
022	Copper (mg / kg (ppm))	0968	76.500	1.0000	75.517	4.5095	1.9532	90	0.22	022.43	0
022	Copper (mg / kg (ppm))	0186	76.500	1.0000	75.517	4.5095	1.9532	90	0.22	022.52	0
022	Copper (mg / kg (ppm))	0229	76.580	1.6600	75.517	4.5095	1.9532	90	0.24	022.41	0
022	Copper (mg / kg (ppm))	0508	76.582	2.1650	75.517	4.5095	1.9532	90	0.24	022.43	0
022	Copper (mg / kg (ppm))	2033	76.850	2.1000	75.517	4.5095	1.9532	90	0.30	022.43	0
022	Copper (mg / kg (ppm))	2012	76.860	0.94000	75.517	4.5095	1.9532	90	0.30	022.41	0
022	Copper (mg / kg (ppm))	0536	76.875	1.6500	75.517	4.5095	1.9532	90	0.30	022.31	0
022	Copper (mg / kg (ppm))	0366	77.000	0.00000	75.517	4.5095	1.9532	90	0.33	022.42	0
022	Copper (mg / kg (ppm))	0208	77.150	1.7000	75.517	4.5095	1.9532	90	0.36	022.31	0
022	Copper (mg / kg (ppm))	2004	77.200	1.2000	75.517	4.5095	1.9532	90	0.37	022.44	0
022	Copper (mg / kg (ppm))	0948	77.237	0.00800	75.517	4.5095	1.9532	90	0.38	022.33	0
022	Copper (mg / kg (ppm))	0726	77.250	0.04000	75.517	4.5095	1.9532	90	0.38	022.42	0
022	Copper (mg / kg (ppm))	0017	77.329	2.3740	75.517	4.5095	1.9532	90	0.40	022.43	0
022	Copper (mg / kg (ppm))	0572	77.450	3.1000	75.517	4.5095	1.9532	90	0.43	022.53	0
022	Copper (mg / kg (ppm))	0175	78.000	0.00000	75.517	4.5095	1.9532	90	0.55	022.31	0
022	Copper (mg / kg (ppm))	0964	78.167	5.2137	75.517	4.5095	1.9532	90	0.59	022.43	0
022	Copper (mg / kg (ppm))	0692	78.350	0.30000	75.517	4.5095	1.9532	90	0.63	022.42	0
022	Copper (mg / kg (ppm))	0675	78.470	4.9600	75.517	4.5095	1.9532	90	0.65	022.43	0
022	Copper (mg / kg (ppm))	2144	78.505	0.53000	75.517	4.5095	1.9532	90	0.66	022.99	0
022	Copper (mg / kg (ppm))	0354	78.650	0.06000	75.517	4.5095	1.9532	90	0.69	022.31	0
022	Copper (mg / kg (ppm))	0413	79.300	2.8000	75.517	4.5095	1.9532	90	0.84	022.53	0
022	Copper (mg / kg (ppm))	0027	79.422	0.40600	75.517	4.5095	1.9532	90	0.87	022.43	0
022	Copper (mg / kg (ppm))	0171	79.430	2.4000	75.517	4.5095	1.9532	90	0.87	022.41	0
022	Copper (mg / kg (ppm))	0512	79.715	3.5900	75.517	4.5095	1.9532	90	0.93	022.43	0
022	Copper (mg / kg (ppm))	0003	80.000	2.0000	75.517	4.5095	1.9532	90	0.99	022.41	0
022	Copper (mg / kg (ppm))	0425	80.200	1.2000	75.517	4.5095	1.9532	90	1.04	022.43	0
022	Copper (mg / kg (ppm))	0083	80.500	3.0000	75.517	4.5095	1.9532	90	1.10	022.43	0
022	Copper (mg / kg (ppm))	0042	80.750	3.1000	75.517	4.5095	1.9532	90	1.16	022.42	0
022	Copper (mg / kg (ppm))	0560	80.750	1.1000	75.517	4.5095	1.9532	90	1.16	022.42	0
022	Copper (mg / kg (ppm))	0190	80.870	1.8400	75.517	4.5095	1.9532	90	1.19	022.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
022	Copper (mg / kg (ppm))	0098	81.295	1.0300	75.517	4.5095	1.9532	90	1.28	022.41	0
022	Copper (mg / kg (ppm))	0413	81.550	4.3000	75.517	4.5095	1.9532	90	1.34	022.42	0
022	Copper (mg / kg (ppm))	0186	82.000	2.0000	75.517	4.5095	1.9532	90	1.44	022.42	0
022	Copper (mg / kg (ppm))	0294	82.100	1.6000	75.517	4.5095	1.9532	90	1.46	022.42	0
022	Copper (mg / kg (ppm))	0353	82.500	5.0000	75.517	4.5095	1.9532	90	1.55	022.43	0
022	Copper (mg / kg (ppm))	2053	82.535	0.87000	75.517	4.5095	1.9532	90	1.56	022.42	0
022	Copper (mg / kg (ppm))	0358	83.335	1.9100	75.517	4.5095	1.9532	90	1.73	022.41	0
022	Copper (mg / kg (ppm))	0226	86.000	6.0000	75.517	4.5095	1.9532	90	2.32	022.41	0
022	Copper (mg / kg (ppm))	0685	92.445	0.45000	75.517	4.5095	1.9532	90	3.75	022.42	0
022	Copper (mg / kg (ppm))	2089	68.725	12.870	75.517	4.5095	1.9532	90	-1.51	022.43	1
022	Copper (mg / kg (ppm))	0035	75.500	11.000	75.517	4.5095	1.9532	90	0.00	022.43	1
022	Copper (mg / kg (ppm))	0011	25.568	5.2050	75.517	4.5095	1.9532	90	-11.08	022.41	2
023	Fluorine (mg / kg (ppm))	0910	51.750	4.5000				1		023.01	0
024	Iodine (mg / kg (ppm))	0171	8.8400	0.30000				1		024.03	0
025	Iron (mg / kg (ppm))	0692	418.00	34.000	586.82	45.670	17.452	85	-3.70	025.42	0
025	Iron (mg / kg (ppm))	0011	436.79	26.310	586.82	45.670	17.452	85	-3.29	025.41	0
025	Iron (mg / kg (ppm))	0511	484.50	5.0000	586.82	45.670	17.452	85	-2.24	025.41	0
025	Iron (mg / kg (ppm))	2022	495.00	2.0000	586.82	45.670	17.452	85	-2.01	025.31	0
025	Iron (mg / kg (ppm))	0186	504.50	17.000	586.82	45.670	17.452	85	-1.80	025.42	0
025	Iron (mg / kg (ppm))	0726	510.90	1.6000	586.82	45.670	17.452	85	-1.66	025.42	0
025	Iron (mg / kg (ppm))	0939	517.05	39.440	586.82	45.670	17.452	85	-1.53	025.31	0
025	Iron (mg / kg (ppm))	0353	520.00	4.0000	586.82	45.670	17.452	85	-1.46	025.43	0
025	Iron (mg / kg (ppm))	0405	523.06	0.34000	586.82	45.670	17.452	85	-1.40	025.41	0
025	Iron (mg / kg (ppm))	0186	533.00	16.000	586.82	45.670	17.452	85	-1.18	025.52	0
025	Iron (mg / kg (ppm))	0345	536.00	8.0000	586.82	45.670	17.452	85	-1.11	025.43	0
025	Iron (mg / kg (ppm))	0520	538.00	14.000	586.82	45.670	17.452	85	-1.07	025.41	0
025	Iron (mg / kg (ppm))	0242	539.50	9.0000	586.82	45.670	17.452	85	-1.04	025.99	0
025	Iron (mg / kg (ppm))	0100	540.50	17.000	586.82	45.670	17.452	85	-1.01	025.99	0
025	Iron (mg / kg (ppm))	0619	542.00	12.000	586.82	45.670	17.452	85	-0.98	025.41	0
025	Iron (mg / kg (ppm))	0154	547.38	49.370	586.82	45.670	17.452	85	-0.86	025.52	0
025	Iron (mg / kg (ppm))	0693	551.97	5.4130	586.82	45.670	17.452	85	-0.76	025.42	0
025	Iron (mg / kg (ppm))	0051	552.50	17.000	586.82	45.670	17.452	85	-0.75	025.41	0
025	Iron (mg / kg (ppm))	0190	553.68	10.140	586.82	45.670	17.452	85	-0.73	025.42	0
025	Iron (mg / kg (ppm))	0018	555.00	70.000	586.82	45.670	17.452	85	-0.70	025.43	0
025	Iron (mg / kg (ppm))	0529	556.70	17.000	586.82	45.670	17.452	85	-0.66	025.31	0
025	Iron (mg / kg (ppm))	0968	557.50	5.0000	586.82	45.670	17.452	85	-0.64	025.43	0
025	Iron (mg / kg (ppm))	0278	558.50	13.000	586.82	45.670	17.452	85	-0.62	025.42	0
025	Iron (mg / kg (ppm))	2012	558.71	9.9700	586.82	45.670	17.452	85	-0.62	025.41	0
025	Iron (mg / kg (ppm))	0629	560.00	8.0000	586.82	45.670	17.452	85	-0.59	025.31	0
025	Iron (mg / kg (ppm))	0263	561.65	9.7000	586.82	45.670	17.452	85	-0.55	025.42	0
025	Iron (mg / kg (ppm))	0354	565.55	0.19000	586.82	45.670	17.452	85	-0.47	025.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
025	Iron (mg / kg (ppm))	0563	568.79	13.790	586.82	45.670	17.452	85	-0.39	025.31	0
025	Iron (mg / kg (ppm))	0208	569.40	31.800	586.82	45.670	17.452	85	-0.38	025.41	0
025	Iron (mg / kg (ppm))	0169	569.50	1.0000	586.82	45.670	17.452	85	-0.38	025.43	0
025	Iron (mg / kg (ppm))	0265	572.00	54.000	586.82	45.670	17.452	85	-0.32	025.42	0
025	Iron (mg / kg (ppm))	2145	572.26	3.7700	586.82	45.670	17.452	85	-0.32	025.41	0
025	Iron (mg / kg (ppm))	0034	574.40	4.2000	586.82	45.670	17.452	85	-0.27	025.43	0
025	Iron (mg / kg (ppm))	0720	576.65	8.1000	586.82	45.670	17.452	85	-0.22	025.41	0
025	Iron (mg / kg (ppm))	0045	577.00	30.000	586.82	45.670	17.452	85	-0.22	025.42	0
025	Iron (mg / kg (ppm))	0049	577.98	12.560	586.82	45.670	17.452	85	-0.19	025.41	0
025	Iron (mg / kg (ppm))	0413	581.50	13.000	586.82	45.670	17.452	85	-0.12	025.42	0
025	Iron (mg / kg (ppm))	0407	582.82	0.31300	586.82	45.670	17.452	85	-0.09	025.41	0
025	Iron (mg / kg (ppm))	0017	584.60	11.719	586.82	45.670	17.452	85	-0.05	025.43	0
025	Iron (mg / kg (ppm))	0425	584.70	0.60000	586.82	45.670	17.452	85	-0.05	025.43	0
025	Iron (mg / kg (ppm))	0918	586.02	28.430	586.82	45.670	17.452	85	-0.02	025.43	0
025	Iron (mg / kg (ppm))	0003	588.50	17.000	586.82	45.670	17.452	85	0.04	025.41	0
025	Iron (mg / kg (ppm))	0955	590.00	68.000	586.82	45.670	17.452	85	0.07	025.41	0
025	Iron (mg / kg (ppm))	0148	590.75	16.500	586.82	45.670	17.452	85	0.09	025.41	0
025	Iron (mg / kg (ppm))	2144	591.33	1.0300	586.82	45.670	17.452	85	0.10	025.99	0
025	Iron (mg / kg (ppm))	0675	591.74	16.110	586.82	45.670	17.452	85	0.11	025.43	0
025	Iron (mg / kg (ppm))	2004	592.00	18.000	586.82	45.670	17.452	85	0.11	025.41	0
025	Iron (mg / kg (ppm))	0175	593.00	30.000	586.82	45.670	17.452	85	0.14	025.31	0
025	Iron (mg / kg (ppm))	0037	593.95	2.7000	586.82	45.670	17.452	85	0.16	025.42	0
025	Iron (mg / kg (ppm))	0598	595.35	1.1000	586.82	45.670	17.452	85	0.19	025.41	0
025	Iron (mg / kg (ppm))	2053	595.50	17.000	586.82	45.670	17.452	85	0.19	025.42	0
025	Iron (mg / kg (ppm))	0171	599.15	0.30000	586.82	45.670	17.452	85	0.27	025.41	0
025	Iron (mg / kg (ppm))	2033	600.00	0.00000	586.82	45.670	17.452	85	0.29	025.43	0
025	Iron (mg / kg (ppm))	2062	601.20	5.8000	586.82	45.670	17.452	85	0.31	025.31	0
025	Iron (mg / kg (ppm))	0294	601.46	2.0800	586.82	45.670	17.452	85	0.32	025.42	0
025	Iron (mg / kg (ppm))	0098	602.25	59.700	586.82	45.670	17.452	85	0.34	025.41	0
025	Iron (mg / kg (ppm))	0019	603.40	5.4000	586.82	45.670	17.452	85	0.36	025.41	0
025	Iron (mg / kg (ppm))	0510	603.50	3.0000	586.82	45.670	17.452	85	0.37	025.43	0
025	Iron (mg / kg (ppm))	0878	607.19	6.2900	586.82	45.670	17.452	85	0.45	025.31	0
025	Iron (mg / kg (ppm))	0689	607.40	21.800	586.82	45.670	17.452	85	0.45	025.31	0
025	Iron (mg / kg (ppm))	0164	608.50	31.000	586.82	45.670	17.452	85	0.47	025.41	0
025	Iron (mg / kg (ppm))	0226	611.50	51.000	586.82	45.670	17.452	85	0.54	025.41	0
025	Iron (mg / kg (ppm))	0366	613.00	8.0000	586.82	45.670	17.452	85	0.57	025.42	0
025	Iron (mg / kg (ppm))	0870	613.95	41.100	586.82	45.670	17.452	85	0.59	025.42	0
025	Iron (mg / kg (ppm))	0042	617.00	28.000	586.82	45.670	17.452	85	0.66	025.42	0
025	Iron (mg / kg (ppm))	0560	618.90	3.4000	586.82	45.670	17.452	85	0.70	025.42	0
025	Iron (mg / kg (ppm))	0083	619.50	11.000	586.82	45.670	17.452	85	0.72	025.43	0
025	Iron (mg / kg (ppm))	0861	620.50	9.0000	586.82	45.670	17.452	85	0.74	025.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
025	Iron (mg / kg (ppm))	2089	623.93	35.300	586.82	45.670	17.452	85	0.81	025.43	0
025	Iron (mg / kg (ppm))	0014	625.50	5.0000	586.82	45.670	17.452	85	0.85	025.31	0
025	Iron (mg / kg (ppm))	0229	629.60	0.80000	586.82	45.670	17.452	85	0.94	025.41	0
025	Iron (mg / kg (ppm))	0964	631.63	10.383	586.82	45.670	17.452	85	0.98	025.43	0
025	Iron (mg / kg (ppm))	0910	632.50	35.000	586.82	45.670	17.452	85	1.00	025.41	0
025	Iron (mg / kg (ppm))	0297	640.00	44.000	586.82	45.670	17.452	85	1.16	025.43	0
025	Iron (mg / kg (ppm))	0098	643.15	93.100	586.82	45.670	17.452	85	1.23	025.43	0
025	Iron (mg / kg (ppm))	0508	643.68	46.150	586.82	45.670	17.452	85	1.24	025.43	0
025	Iron (mg / kg (ppm))	0208	650.00	2.0000	586.82	45.670	17.452	85	1.38	025.31	0
025	Iron (mg / kg (ppm))	0512	651.10	0.40000	586.82	45.670	17.452	85	1.41	025.43	0
025	Iron (mg / kg (ppm))	0682	651.47	0.00000	586.82	45.670	17.452	85	1.42	025.43	0
025	Iron (mg / kg (ppm))	0553	653.00	18.000	586.82	45.670	17.452	85	1.45	025.53	0
025	Iron (mg / kg (ppm))	0358	666.12	21.890	586.82	45.670	17.452	85	1.74	025.41	0
025	Iron (mg / kg (ppm))	0413	667.50	11.000	586.82	45.670	17.452	85	1.77	025.53	0
025	Iron (mg / kg (ppm))	0035	685.50	49.000	586.82	45.670	17.452	85	2.16	025.42	0
025	Iron (mg / kg (ppm))	0948	747.77	0.35600	586.82	45.670	17.452	85	3.52	025.33	0
025	Iron (mg / kg (ppm))	0035	773.50	3.0000	586.82	45.670	17.452	85	4.09	025.43	0
025	Iron (mg / kg (ppm))	0874	745.50	199.00	586.82	45.670	17.452	85	3.47	025.31	1
025	Iron (mg / kg (ppm))	0536	147.13	1.0500	586.82	45.670	17.452	85	-9.63	025.31	2
027	Magnesium (%)	0536	0.08000	0.00000	0.16068	0.01104	0.00467	99	-7.31	027.31	0
027	Magnesium (%)	0889	0.10835	0.00950	0.16068	0.01104	0.00467	99	-4.74	027.99	0
027	Magnesium (%)	2129	0.13555	0.00850	0.16068	0.01104	0.00467	99	-2.28	027.42	0
027	Magnesium (%)	0169	0.14000	0.00000	0.16068	0.01104	0.00467	99	-1.87	027.32	0
027	Magnesium (%)	0675	0.14000	0.02000	0.16068	0.01104	0.00467	99	-1.87	027.43	0
027	Magnesium (%)	0407	0.14075	0.00050	0.16068	0.01104	0.00467	99	-1.81	027.41	0
027	Magnesium (%)	0345	0.14100	0.00200	0.16068	0.01104	0.00467	99	-1.78	027.43	0
027	Magnesium (%)	0148	0.14500	0.01000	0.16068	0.01104	0.00467	99	-1.42	027.41	0
027	Magnesium (%)	0511	0.14500	0.01000	0.16068	0.01104	0.00467	99	-1.42	027.41	0
027	Magnesium (%)	0045	0.14650	0.00500	0.16068	0.01104	0.00467	99	-1.28	027.42	0
027	Magnesium (%)	0650	0.14800	0.01200	0.16068	0.01104	0.00467	99	-1.15	027.31	0
027	Magnesium (%)	2144	0.14820	0.00040	0.16068	0.01104	0.00467	99	-1.13	027.99	0
027	Magnesium (%)	0011	0.14885	0.00070	0.16068	0.01104	0.00467	99	-1.07	027.41	0
027	Magnesium (%)	0939	0.15000	0.00000	0.16068	0.01104	0.00467	99	-0.97	027.31	0
027	Magnesium (%)	0720	0.15000	0.00000	0.16068	0.01104	0.00467	99	-0.97	027.41	0
027	Magnesium (%)	2012	0.15000	0.00000	0.16068	0.01104	0.00467	99	-0.97	027.41	0
027	Magnesium (%)	0242	0.15000	0.00000	0.16068	0.01104	0.00467	99	-0.97	027.99	0
027	Magnesium (%)	0689	0.15000	0.02000	0.16068	0.01104	0.00467	99	-0.97	027.31	0
027	Magnesium (%)	0229	0.15150	0.00500	0.16068	0.01104	0.00467	99	-0.83	027.41	0
027	Magnesium (%)	0164	0.15200	0.00200	0.16068	0.01104	0.00467	99	-0.79	027.41	0
027	Magnesium (%)	0918	0.15200	0.00600	0.16068	0.01104	0.00467	99	-0.79	027.43	0
027	Magnesium (%)	0051	0.15315	0.01010	0.16068	0.01104	0.00467	99	-0.68	027.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
027	Magnesium (%)	0563	0.15320	0.00340	0.16068	0.01104	0.00467	99	-0.68	027.31	0
027	Magnesium (%)	0175	0.15500	0.01000	0.16068	0.01104	0.00467	99	-0.51	027.31	0
027	Magnesium (%)	0003	0.15500	0.01000	0.16068	0.01104	0.00467	99	-0.51	027.41	0
027	Magnesium (%)	0520	0.15500	0.01000	0.16068	0.01104	0.00467	99	-0.51	027.41	0
027	Magnesium (%)	0366	0.15500	0.01000	0.16068	0.01104	0.00467	99	-0.51	027.42	0
027	Magnesium (%)	0692	0.15500	0.01000	0.16068	0.01104	0.00467	99	-0.51	027.42	0
027	Magnesium (%)	0083	0.15500	0.01000	0.16068	0.01104	0.00467	99	-0.51	027.43	0
027	Magnesium (%)	0510	0.15500	0.01000	0.16068	0.01104	0.00467	99	-0.51	027.43	0
027	Magnesium (%)	0508	0.15510	0.00380	0.16068	0.01104	0.00467	99	-0.51	027.43	0
027	Magnesium (%)	0042	0.15550	0.00300	0.16068	0.01104	0.00467	99	-0.47	027.42	0
027	Magnesium (%)	0955	0.15550	0.00100	0.16068	0.01104	0.00467	99	-0.47	027.44	0
027	Magnesium (%)	0154	0.15585	0.00490	0.16068	0.01104	0.00467	99	-0.44	027.52	0
027	Magnesium (%)	0017	0.15600	0.00000	0.16068	0.01104	0.00467	99	-0.42	027.43	0
027	Magnesium (%)	0043	0.15600	0.00200	0.16068	0.01104	0.00467	99	-0.42	027.43	0
027	Magnesium (%)	0202	0.15675	0.00170	0.16068	0.01104	0.00467	99	-0.36	027.43	0
027	Magnesium (%)	0870	0.15710	0.01020	0.16068	0.01104	0.00467	99	-0.32	027.42	0
027	Magnesium (%)	0560	0.15725	0.00150	0.16068	0.01104	0.00467	99	-0.31	027.42	0
027	Magnesium (%)	0034	0.15785	0.00650	0.16068	0.01104	0.00467	99	-0.26	027.43	0
027	Magnesium (%)	0861	0.15800	0.00000	0.16068	0.01104	0.00467	99	-0.24	027.43	0
027	Magnesium (%)	2022	0.15900	0.00200	0.16068	0.01104	0.00467	99	-0.15	027.31	0
027	Magnesium (%)	0629	0.16000	0.00000	0.16068	0.01104	0.00467	99	-0.06	027.31	0
027	Magnesium (%)	0878	0.16000	0.00000	0.16068	0.01104	0.00467	99	-0.06	027.31	0
027	Magnesium (%)	0049	0.16000	0.00000	0.16068	0.01104	0.00467	99	-0.06	027.41	0
027	Magnesium (%)	0139	0.16000	0.00300	0.16068	0.01104	0.00467	99	-0.06	027.41	0
027	Magnesium (%)	0413	0.16000	0.00000	0.16068	0.01104	0.00467	99	-0.06	027.41	0
027	Magnesium (%)	0910	0.16000	0.02000	0.16068	0.01104	0.00467	99	-0.06	027.41	0
027	Magnesium (%)	0265	0.16000	0.00000	0.16068	0.01104	0.00467	99	-0.06	027.42	0
027	Magnesium (%)	0294	0.16000	0.00000	0.16068	0.01104	0.00467	99	-0.06	027.42	0
027	Magnesium (%)	0297	0.16000	0.00000	0.16068	0.01104	0.00467	99	-0.06	027.43	0
027	Magnesium (%)	2089	0.16000	0.00000	0.16068	0.01104	0.00467	99	-0.06	027.43	0
027	Magnesium (%)	2004	0.16000	0.00200	0.16068	0.01104	0.00467	99	-0.06	027.44	0
027	Magnesium (%)	0553	0.16000	0.00000	0.16068	0.01104	0.00467	99	-0.06	027.53	0
027	Magnesium (%)	0590	0.16000	0.00000	0.16068	0.01104	0.00467	99	-0.06	027.99	0
027	Magnesium (%)	0964	0.16145	0.01190	0.16068	0.01104	0.00467	99	0.07	027.43	0
027	Magnesium (%)	0171	0.16150	0.00100	0.16068	0.01104	0.00467	99	0.07	027.41	0
027	Magnesium (%)	0142	0.16200	0.00400	0.16068	0.01104	0.00467	99	0.12	027.31	0
027	Magnesium (%)	0529	0.16200	0.00400	0.16068	0.01104	0.00467	99	0.12	027.31	0
027	Magnesium (%)	0874	0.16250	0.00300	0.16068	0.01104	0.00467	99	0.16	027.31	0
027	Magnesium (%)	0208	0.16300	0.00600	0.16068	0.01104	0.00467	99	0.21	027.41	0
027	Magnesium (%)	0263	0.16355	0.00210	0.16068	0.01104	0.00467	99	0.26	027.42	0
027	Magnesium (%)	0512	0.16390	0.00220	0.16068	0.01104	0.00467	99	0.29	027.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
027	Magnesium (%)	0948	0.16450	0.00100	0.16068	0.01104	0.00467	99	0.35	027.33	0
027	Magnesium (%)	2145	0.16450	0.00100	0.16068	0.01104	0.00467	99	0.35	027.41	0
027	Magnesium (%)	0186	0.16450	0.00500	0.16068	0.01104	0.00467	99	0.35	027.52	0
027	Magnesium (%)	0019	0.16500	0.01000	0.16068	0.01104	0.00467	99	0.39	027.41	0
027	Magnesium (%)	0035	0.16500	0.01000	0.16068	0.01104	0.00467	99	0.39	027.42	0
027	Magnesium (%)	2053	0.16500	0.01000	0.16068	0.01104	0.00467	99	0.39	027.42	0
027	Magnesium (%)	0027	0.16500	0.00200	0.16068	0.01104	0.00467	99	0.39	027.43	0
027	Magnesium (%)	0098	0.16500	0.01000	0.16068	0.01104	0.00467	99	0.39	027.43	0
027	Magnesium (%)	0425	0.16500	0.01000	0.16068	0.01104	0.00467	99	0.39	027.43	0
027	Magnesium (%)	0413	0.16500	0.01000	0.16068	0.01104	0.00467	99	0.39	027.53	0
027	Magnesium (%)	0208	0.16700	0.00000	0.16068	0.01104	0.00467	99	0.57	027.31	0
027	Magnesium (%)	0357	0.16730	0.00120	0.16068	0.01104	0.00467	99	0.60	027.42	0
027	Magnesium (%)	0572	0.16750	0.00100	0.16068	0.01104	0.00467	99	0.62	027.53	0
027	Magnesium (%)	0693	0.16800	0.00200	0.16068	0.01104	0.00467	99	0.66	027.42	0
027	Magnesium (%)	0226	0.17000	0.00000	0.16068	0.01104	0.00467	99	0.84	027.41	0
027	Magnesium (%)	0037	0.17000	0.00000	0.16068	0.01104	0.00467	99	0.84	027.42	0
027	Magnesium (%)	0035	0.17000	0.00000	0.16068	0.01104	0.00467	99	0.84	027.43	0
027	Magnesium (%)	2033	0.17000	0.00000	0.16068	0.01104	0.00467	99	0.84	027.43	0
027	Magnesium (%)	0014	0.17050	0.01900	0.16068	0.01104	0.00467	99	0.89	027.31	0
027	Magnesium (%)	0186	0.17100	0.01200	0.16068	0.01104	0.00467	99	0.94	027.42	0
027	Magnesium (%)	0726	0.17395	0.00070	0.16068	0.01104	0.00467	99	1.20	027.42	0
027	Magnesium (%)	0098	0.17500	0.01000	0.16068	0.01104	0.00467	99	1.30	027.41	0
027	Magnesium (%)	0405	0.17500	0.01000	0.16068	0.01104	0.00467	99	1.30	027.41	0
027	Magnesium (%)	0190	0.17500	0.01000	0.16068	0.01104	0.00467	99	1.30	027.42	0
027	Magnesium (%)	0353	0.17500	0.01000	0.16068	0.01104	0.00467	99	1.30	027.43	0
027	Magnesium (%)	0619	0.17550	0.00300	0.16068	0.01104	0.00467	99	1.34	027.41	0
027	Magnesium (%)	0598	0.17590	0.00080	0.16068	0.01104	0.00467	99	1.38	027.41	0
027	Magnesium (%)	0358	0.18000	0.00000	0.16068	0.01104	0.00467	99	1.75	027.41	0
027	Magnesium (%)	0278	0.18000	0.00000	0.16068	0.01104	0.00467	99	1.75	027.42	0
027	Magnesium (%)	0685	0.18000	0.00000	0.16068	0.01104	0.00467	99	1.75	027.42	0
027	Magnesium (%)	0089	0.18000	0.00000	0.16068	0.01104	0.00467	99	1.75	027.43	0
027	Magnesium (%)	0682	0.18000	0.00000	0.16068	0.01104	0.00467	99	1.75	027.43	0
027	Magnesium (%)	0100	0.18000	0.00000	0.16068	0.01104	0.00467	99	1.75	027.99	0
027	Magnesium (%)	0018	0.18200	0.01200	0.16068	0.01104	0.00467	99	1.93	027.43	0
027	Magnesium (%)	0968	0.18550	0.01100	0.16068	0.01104	0.00467	99	2.25	027.43	0
027	Magnesium (%)	0354	0.19000	0.00000	0.16068	0.01104	0.00467	99	2.66	027.31	0
027	Magnesium (%)	2097	0.33515	0.00510	0.16068	0.01104	0.00467	99	15.81	027.31	2
028	Manganese (mg / kg (ppm))	0689	78.850	1.1000	99.935	5.8796	3.0228	75	-3.59	028.31	0
028	Manganese (mg / kg (ppm))	0536	84.125	0.13000	99.935	5.8796	3.0228	75	-2.69	028.31	0
028	Manganese (mg / kg (ppm))	0948	86.454	0.00500	99.935	5.8796	3.0228	75	-2.29	028.33	0
028	Manganese (mg / kg (ppm))	0169	90.000	1.8000	99.935	5.8796	3.0228	75	-1.69	028.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
028	Manganese (mg / kg (ppm))	0675	92.170	12.440	99.935	5.8796	3.0228	75	-1.32	028.43	0
028	Manganese (mg / kg (ppm))	0720	93.180	2.2000	99.935	5.8796	3.0228	75	-1.15	028.41	0
028	Manganese (mg / kg (ppm))	0870	93.580	1.7000	99.935	5.8796	3.0228	75	-1.08	028.42	0
028	Manganese (mg / kg (ppm))	0037	93.600	16.000	99.935	5.8796	3.0228	75	-1.08	028.42	0
028	Manganese (mg / kg (ppm))	0861	93.600	2.2000	99.935	5.8796	3.0228	75	-1.08	028.43	0
028	Manganese (mg / kg (ppm))	0263	93.850	0.90000	99.935	5.8796	3.0228	75	-1.03	028.42	0
028	Manganese (mg / kg (ppm))	0939	94.145	2.0700	99.935	5.8796	3.0228	75	-0.98	028.31	0
028	Manganese (mg / kg (ppm))	2012	94.180	3.2800	99.935	5.8796	3.0228	75	-0.98	028.41	0
028	Manganese (mg / kg (ppm))	0354	94.215	0.75000	99.935	5.8796	3.0228	75	-0.97	028.31	0
028	Manganese (mg / kg (ppm))	0164	94.500	5.0000	99.935	5.8796	3.0228	75	-0.92	028.41	0
028	Manganese (mg / kg (ppm))	0353	94.500	3.0000	99.935	5.8796	3.0228	75	-0.92	028.43	0
028	Manganese (mg / kg (ppm))	0051	94.805	2.4100	99.935	5.8796	3.0228	75	-0.87	028.41	0
028	Manganese (mg / kg (ppm))	0693	94.991	4.3010	99.935	5.8796	3.0228	75	-0.84	028.42	0
028	Manganese (mg / kg (ppm))	0265	95.000	6.0000	99.935	5.8796	3.0228	75	-0.84	028.42	0
028	Manganese (mg / kg (ppm))	0017	95.101	0.52700	99.935	5.8796	3.0228	75	-0.82	028.43	0
028	Manganese (mg / kg (ppm))	0520	95.500	3.0000	99.935	5.8796	3.0228	75	-0.75	028.41	0
028	Manganese (mg / kg (ppm))	0148	95.550	5.3000	99.935	5.8796	3.0228	75	-0.75	028.41	0
028	Manganese (mg / kg (ppm))	0208	95.570	0.42000	99.935	5.8796	3.0228	75	-0.74	028.41	0
028	Manganese (mg / kg (ppm))	0034	96.385	3.1100	99.935	5.8796	3.0228	75	-0.60	028.43	0
028	Manganese (mg / kg (ppm))	0878	96.740	0.84000	99.935	5.8796	3.0228	75	-0.54	028.31	0
028	Manganese (mg / kg (ppm))	0043	96.800	0.20000	99.935	5.8796	3.0228	75	-0.53	028.43	0
028	Manganese (mg / kg (ppm))	0345	97.000	0.40000	99.935	5.8796	3.0228	75	-0.50	028.43	0
028	Manganese (mg / kg (ppm))	0407	97.127	0.97660	99.935	5.8796	3.0228	75	-0.48	028.41	0
028	Manganese (mg / kg (ppm))	0425	97.750	1.1000	99.935	5.8796	3.0228	75	-0.37	028.43	0
028	Manganese (mg / kg (ppm))	0968	98.000	2.0000	99.935	5.8796	3.0228	75	-0.33	028.43	0
028	Manganese (mg / kg (ppm))	0171	98.180	0.68000	99.935	5.8796	3.0228	75	-0.30	028.41	0
028	Manganese (mg / kg (ppm))	2062	98.450	1.3000	99.935	5.8796	3.0228	75	-0.25	028.31	0
028	Manganese (mg / kg (ppm))	2033	98.925	0.15000	99.935	5.8796	3.0228	75	-0.17	028.43	0
028	Manganese (mg / kg (ppm))	0278	99.000	2.0000	99.935	5.8796	3.0228	75	-0.16	028.42	0
028	Manganese (mg / kg (ppm))	0413	99.000	2.0000	99.935	5.8796	3.0228	75	-0.16	028.42	0
028	Manganese (mg / kg (ppm))	0510	99.000	0.00000	99.935	5.8796	3.0228	75	-0.16	028.43	0
028	Manganese (mg / kg (ppm))	2004	99.000	0.40000	99.935	5.8796	3.0228	75	-0.16	028.44	0
028	Manganese (mg / kg (ppm))	0918	99.019	0.48770	99.935	5.8796	3.0228	75	-0.16	028.43	0
028	Manganese (mg / kg (ppm))	0045	99.250	1.5000	99.935	5.8796	3.0228	75	-0.12	028.42	0
028	Manganese (mg / kg (ppm))	0560	100.25	0.90000	99.935	5.8796	3.0228	75	0.05	028.42	0
028	Manganese (mg / kg (ppm))	0685	100.46	1.4200	99.935	5.8796	3.0228	75	0.09	028.42	0
028	Manganese (mg / kg (ppm))	0242	100.50	3.0000	99.935	5.8796	3.0228	75	0.10	028.99	0
028	Manganese (mg / kg (ppm))	0563	100.69	4.5296	99.935	5.8796	3.0228	75	0.13	028.31	0
028	Manganese (mg / kg (ppm))	0049	100.75	6.5300	99.935	5.8796	3.0228	75	0.14	028.41	0
028	Manganese (mg / kg (ppm))	0964	100.80	6.1807	99.935	5.8796	3.0228	75	0.15	028.43	0
028	Manganese (mg / kg (ppm))	0229	101.20	0.60000	99.935	5.8796	3.0228	75	0.22	028.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
028	Manganese (mg / kg (ppm))	0508	101.63	3.7760	99.935	5.8796	3.0228	75	0.29	028.43	0
028	Manganese (mg / kg (ppm))	0208	102.00	10.000	99.935	5.8796	3.0228	75	0.35	028.31	0
028	Manganese (mg / kg (ppm))	0083	102.00	2.0000	99.935	5.8796	3.0228	75	0.35	028.43	0
028	Manganese (mg / kg (ppm))	0590	102.00	0.00000	99.935	5.8796	3.0228	75	0.35	028.99	0
028	Manganese (mg / kg (ppm))	2053	102.50	1.0000	99.935	5.8796	3.0228	75	0.44	028.42	0
028	Manganese (mg / kg (ppm))	0098	102.98	6.6400	99.935	5.8796	3.0228	75	0.52	028.43	0
028	Manganese (mg / kg (ppm))	0572	103.00	2.0000	99.935	5.8796	3.0228	75	0.52	028.53	0
028	Manganese (mg / kg (ppm))	0553	103.30	7.4000	99.935	5.8796	3.0228	75	0.57	028.53	0
028	Manganese (mg / kg (ppm))	0035	103.50	15.000	99.935	5.8796	3.0228	75	0.61	028.42	0
028	Manganese (mg / kg (ppm))	0413	103.50	1.0000	99.935	5.8796	3.0228	75	0.61	028.53	0
028	Manganese (mg / kg (ppm))	0692	104.00	0.00000	99.935	5.8796	3.0228	75	0.69	028.42	0
028	Manganese (mg / kg (ppm))	0019	104.15	2.5000	99.935	5.8796	3.0228	75	0.72	028.41	0
028	Manganese (mg / kg (ppm))	0910	104.50	1.0000	99.935	5.8796	3.0228	75	0.78	028.41	0
028	Manganese (mg / kg (ppm))	0035	104.50	21.000	99.935	5.8796	3.0228	75	0.78	028.43	0
028	Manganese (mg / kg (ppm))	0186	104.50	3.0000	99.935	5.8796	3.0228	75	0.78	028.52	0
028	Manganese (mg / kg (ppm))	0297	105.00	6.0000	99.935	5.8796	3.0228	75	0.86	028.43	0
028	Manganese (mg / kg (ppm))	0682	105.55	0.00000	99.935	5.8796	3.0228	75	0.95	028.43	0
028	Manganese (mg / kg (ppm))	0529	105.85	0.50000	99.935	5.8796	3.0228	75	1.01	028.31	0
028	Manganese (mg / kg (ppm))	0098	105.98	12.640	99.935	5.8796	3.0228	75	1.03	028.41	0
028	Manganese (mg / kg (ppm))	2089	106.26	0.19000	99.935	5.8796	3.0228	75	1.07	028.43	0
028	Manganese (mg / kg (ppm))	0366	106.50	3.0000	99.935	5.8796	3.0228	75	1.12	028.42	0
028	Manganese (mg / kg (ppm))	0100	107.50	1.0000	99.935	5.8796	3.0228	75	1.29	028.99	0
028	Manganese (mg / kg (ppm))	0598	107.55	3.3000	99.935	5.8796	3.0228	75	1.30	028.41	0
028	Manganese (mg / kg (ppm))	0202	107.58	0.20200	99.935	5.8796	3.0228	75	1.30	028.43	0
028	Manganese (mg / kg (ppm))	0186	109.00	4.0000	99.935	5.8796	3.0228	75	1.54	028.42	0
028	Manganese (mg / kg (ppm))	0357	111.11	0.25200	99.935	5.8796	3.0228	75	1.90	028.42	0
028	Manganese (mg / kg (ppm))	0175	113.50	3.0000	99.935	5.8796	3.0228	75	2.31	028.31	0
028	Manganese (mg / kg (ppm))	0358	114.86	0.64000	99.935	5.8796	3.0228	75	2.54	028.41	0
028	Manganese (mg / kg (ppm))	0294	118.92	0.56000	99.935	5.8796	3.0228	75	3.23	028.42	0
028	Manganese (mg / kg (ppm))	0405	121.68	0.27000	99.935	5.8796	3.0228	75	3.70	028.41	0
028	Manganese (mg / kg (ppm))	0226	114.50	31.000	99.935	5.8796	3.0228	75	2.48	028.41	1
028	Manganese (mg / kg (ppm))	0042	119.00	34.000	99.935	5.8796	3.0228	75	3.24	028.42	1
031	Phosphorus (%)	0889	1.2046	0.00480	1.7409	0.07239	0.03168	134	-7.41	031.99	0
031	Phosphorus (%)	2105	1.2900	0.00000	1.7409	0.07239	0.03168	134	-6.23	031.99	0
031	Phosphorus (%)	0885	1.3165	0.00300	1.7409	0.07239	0.03168	134	-5.86	031.01	0
031	Phosphorus (%)	0536	1.3400	0.02000	1.7409	0.07239	0.03168	134	-5.54	031.06	0
031	Phosphorus (%)	0520	1.4600	0.08000	1.7409	0.07239	0.03168	134	-3.88	031.41	0
031	Phosphorus (%)	0685	1.4950	0.03000	1.7409	0.07239	0.03168	134	-3.40	031.41	0
031	Phosphorus (%)	0186	1.5200	0.02000	1.7409	0.07239	0.03168	134	-3.05	031.42	0
031	Phosphorus (%)	0650	1.5550	0.05000	1.7409	0.07239	0.03168	134	-2.57	031.01	0
031	Phosphorus (%)	0089	1.5900	0.00000	1.7409	0.07239	0.03168	134	-2.08	031.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	0148	1.5950	0.01000	1.7409	0.07239	0.03168	134	-2.02	031.41	0
031	Phosphorus (%)	0874	1.6020	0.02000	1.7409	0.07239	0.03168	134	-1.92	031.01	0
031	Phosphorus (%)	0674	1.6200	0.00000	1.7409	0.07239	0.03168	134	-1.67	031.01	0
031	Phosphorus (%)	0233	1.6250	0.01000	1.7409	0.07239	0.03168	134	-1.60	031.01	0
031	Phosphorus (%)	0345	1.6300	0.04000	1.7409	0.07239	0.03168	134	-1.53	031.43	0
031	Phosphorus (%)	0623	1.6320	0.03000	1.7409	0.07239	0.03168	134	-1.50	031.01	0
031	Phosphorus (%)	2009	1.6384	0.00020	1.7409	0.07239	0.03168	134	-1.42	031.01	0
031	Phosphorus (%)	2129	1.6500	0.10000	1.7409	0.07239	0.03168	134	-1.26	031.42	0
031	Phosphorus (%)	0405	1.6550	0.05000	1.7409	0.07239	0.03168	134	-1.19	031.41	0
031	Phosphorus (%)	0154	1.6571	0.00720	1.7409	0.07239	0.03168	134	-1.16	031.52	0
031	Phosphorus (%)	0169	1.6600	0.02000	1.7409	0.07239	0.03168	134	-1.12	031.03	0
031	Phosphorus (%)	0034	1.6616	0.06990	1.7409	0.07239	0.03168	134	-1.10	031.43	0
031	Phosphorus (%)	0622	1.6686	0.00060	1.7409	0.07239	0.03168	134	-1.00	031.01	0
031	Phosphorus (%)	0144	1.6700	0.08000	1.7409	0.07239	0.03168	134	-0.98	031.41	0
031	Phosphorus (%)	0045	1.6700	0.04000	1.7409	0.07239	0.03168	134	-0.98	031.42	0
031	Phosphorus (%)	0294	1.6700	0.00000	1.7409	0.07239	0.03168	134	-0.98	031.42	0
031	Phosphorus (%)	0511	1.6750	0.09000	1.7409	0.07239	0.03168	134	-0.91	031.01	0
031	Phosphorus (%)	0940	1.6750	0.05000	1.7409	0.07239	0.03168	134	-0.91	031.01	0
031	Phosphorus (%)	0003	1.6750	0.01000	1.7409	0.07239	0.03168	134	-0.91	031.41	0
031	Phosphorus (%)	0676	1.6800	0.02000	1.7409	0.07239	0.03168	134	-0.84	031.99	0
031	Phosphorus (%)	0852	1.6800	0.04000	1.7409	0.07239	0.03168	134	-0.84	031.99	0
031	Phosphorus (%)	0689	1.6900	0.04000	1.7409	0.07239	0.03168	134	-0.70	031.01	0
031	Phosphorus (%)	0100	1.6900	0.02000	1.7409	0.07239	0.03168	134	-0.70	031.99	0
031	Phosphorus (%)	0918	1.6955	0.00300	1.7409	0.07239	0.03168	134	-0.63	031.43	0
031	Phosphorus (%)	0968	1.6970	0.00600	1.7409	0.07239	0.03168	134	-0.61	031.43	0
031	Phosphorus (%)	0693	1.6990	0.00400	1.7409	0.07239	0.03168	134	-0.58	031.42	0
031	Phosphorus (%)	0897	1.7050	0.01000	1.7409	0.07239	0.03168	134	-0.50	031.01	0
031	Phosphorus (%)	0720	1.7050	0.01000	1.7409	0.07239	0.03168	134	-0.50	031.41	0
031	Phosphorus (%)	0186	1.7050	0.09000	1.7409	0.07239	0.03168	134	-0.50	031.52	0
031	Phosphorus (%)	0870	1.7075	0.05300	1.7409	0.07239	0.03168	134	-0.46	031.42	0
031	Phosphorus (%)	0142	1.7082	0.00160	1.7409	0.07239	0.03168	134	-0.45	031.01	0
031	Phosphorus (%)	0123	1.7100	0.04000	1.7409	0.07239	0.03168	134	-0.43	031.41	0
031	Phosphorus (%)	0019	1.7150	0.03000	1.7409	0.07239	0.03168	134	-0.36	031.41	0
031	Phosphorus (%)	0035	1.7150	0.11000	1.7409	0.07239	0.03168	134	-0.36	031.42	0
031	Phosphorus (%)	0278	1.7150	0.01000	1.7409	0.07239	0.03168	134	-0.36	031.42	0
031	Phosphorus (%)	0878	1.7200	0.00000	1.7409	0.07239	0.03168	134	-0.29	031.01	0
031	Phosphorus (%)	2145	1.7200	0.00000	1.7409	0.07239	0.03168	134	-0.29	031.41	0
031	Phosphorus (%)	0037	1.7200	0.02000	1.7409	0.07239	0.03168	134	-0.29	031.42	0
031	Phosphorus (%)	0407	1.7213	0.00040	1.7409	0.07239	0.03168	134	-0.27	031.41	0
031	Phosphorus (%)	0590	1.7250	0.05000	1.7409	0.07239	0.03168	134	-0.22	031.01	0
031	Phosphorus (%)	0903	1.7250	0.01000	1.7409	0.07239	0.03168	134	-0.22	031.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	0042	1.7250	0.15000	1.7409	0.07239	0.03168	134	-0.22	031.42	0
031	Phosphorus (%)	0651	1.7260	0.03200	1.7409	0.07239	0.03168	134	-0.21	031.01	0
031	Phosphorus (%)	0175	1.7300	0.04000	1.7409	0.07239	0.03168	134	-0.15	031.01	0
031	Phosphorus (%)	2012	1.7300	0.04000	1.7409	0.07239	0.03168	134	-0.15	031.41	0
031	Phosphorus (%)	0011	1.7316	0.01260	1.7409	0.07239	0.03168	134	-0.13	031.41	0
031	Phosphorus (%)	0560	1.7318	0.02910	1.7409	0.07239	0.03168	134	-0.13	031.42	0
031	Phosphorus (%)	0139	1.7325	0.01500	1.7409	0.07239	0.03168	134	-0.12	031.41	0
031	Phosphorus (%)	0171	1.7325	0.00500	1.7409	0.07239	0.03168	134	-0.12	031.41	0
031	Phosphorus (%)	0051	1.7349	0.02260	1.7409	0.07239	0.03168	134	-0.08	031.41	0
031	Phosphorus (%)	0683	1.7350	0.01000	1.7409	0.07239	0.03168	134	-0.08	031.01	0
031	Phosphorus (%)	0884	1.7350	0.05000	1.7409	0.07239	0.03168	134	-0.08	031.01	0
031	Phosphorus (%)	0955	1.7350	0.03000	1.7409	0.07239	0.03168	134	-0.08	031.44	0
031	Phosphorus (%)	0629	1.7450	0.01000	1.7409	0.07239	0.03168	134	0.06	031.01	0
031	Phosphorus (%)	0653	1.7490	0.04400	1.7409	0.07239	0.03168	134	0.11	031.01	0
031	Phosphorus (%)	0366	1.7500	0.04000	1.7409	0.07239	0.03168	134	0.13	031.42	0
031	Phosphorus (%)	0297	1.7500	0.04000	1.7409	0.07239	0.03168	134	0.13	031.43	0
031	Phosphorus (%)	2006	1.7510	0.04200	1.7409	0.07239	0.03168	134	0.14	031.01	0
031	Phosphorus (%)	0722	1.7512	0.00610	1.7409	0.07239	0.03168	134	0.14	031.01	0
031	Phosphorus (%)	0263	1.7531	0.00620	1.7409	0.07239	0.03168	134	0.17	031.42	0
031	Phosphorus (%)	0354	1.7550	0.01000	1.7409	0.07239	0.03168	134	0.19	031.01	0
031	Phosphorus (%)	0939	1.7550	0.01000	1.7409	0.07239	0.03168	134	0.19	031.01	0
031	Phosphorus (%)	0049	1.7550	0.01000	1.7409	0.07239	0.03168	134	0.19	031.41	0
031	Phosphorus (%)	0164	1.7550	0.01000	1.7409	0.07239	0.03168	134	0.19	031.41	0
031	Phosphorus (%)	0861	1.7550	0.07000	1.7409	0.07239	0.03168	134	0.19	031.43	0
031	Phosphorus (%)	0563	1.7576	0.03750	1.7409	0.07239	0.03168	134	0.23	031.01	0
031	Phosphorus (%)	2062	1.7596	0.04230	1.7409	0.07239	0.03168	134	0.26	031.01	0
031	Phosphorus (%)	2068	1.7600	0.00000	1.7409	0.07239	0.03168	134	0.26	031.01	0
031	Phosphorus (%)	0619	1.7600	0.00000	1.7409	0.07239	0.03168	134	0.26	031.41	0
031	Phosphorus (%)	0017	1.7600	0.06800	1.7409	0.07239	0.03168	134	0.26	031.43	0
031	Phosphorus (%)	0682	1.7600	0.00000	1.7409	0.07239	0.03168	134	0.26	031.43	0
031	Phosphorus (%)	2089	1.7600	0.00000	1.7409	0.07239	0.03168	134	0.26	031.43	0
031	Phosphorus (%)	2144	1.7600	0.02000	1.7409	0.07239	0.03168	134	0.26	031.99	0
031	Phosphorus (%)	0006	1.7645	0.03300	1.7409	0.07239	0.03168	134	0.33	031.43	0
031	Phosphorus (%)	0895	1.7650	0.01000	1.7409	0.07239	0.03168	134	0.33	031.00	0
031	Phosphorus (%)	0552	1.7650	0.07000	1.7409	0.07239	0.03168	134	0.33	031.01	0
031	Phosphorus (%)	0413	1.7650	0.05000	1.7409	0.07239	0.03168	134	0.33	031.41	0
031	Phosphorus (%)	0508	1.7657	0.05100	1.7409	0.07239	0.03168	134	0.34	031.43	0
031	Phosphorus (%)	0626	1.7660	0.00400	1.7409	0.07239	0.03168	134	0.35	031.01	0
031	Phosphorus (%)	0208	1.7680	0.13400	1.7409	0.07239	0.03168	134	0.37	031.41	0
031	Phosphorus (%)	0035	1.7700	0.02000	1.7409	0.07239	0.03168	134	0.40	031.01	0
031	Phosphorus (%)	0529	1.7700	0.02000	1.7409	0.07239	0.03168	134	0.40	031.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	0298	1.7700	0.02000	1.7409	0.07239	0.03168	134	0.40	031.41	0
031	Phosphorus (%)	0014	1.7700	0.08000	1.7409	0.07239	0.03168	134	0.40	031.42	0
031	Phosphorus (%)	0033	1.7700	0.04000	1.7409	0.07239	0.03168	134	0.40	031.43	0
031	Phosphorus (%)	0675	1.7700	0.10000	1.7409	0.07239	0.03168	134	0.40	031.43	0
031	Phosphorus (%)	0849	1.7715	0.00300	1.7409	0.07239	0.03168	134	0.42	031.01	0
031	Phosphorus (%)	0202	1.7740	0.02800	1.7409	0.07239	0.03168	134	0.46	031.43	0
031	Phosphorus (%)	2022	1.7750	0.01000	1.7409	0.07239	0.03168	134	0.47	031.01	0
031	Phosphorus (%)	0098	1.7820	0.05600	1.7409	0.07239	0.03168	134	0.57	031.41	0
031	Phosphorus (%)	0948	1.7845	0.00100	1.7409	0.07239	0.03168	134	0.60	031.01	0
031	Phosphorus (%)	0357	1.7850	0.00000	1.7409	0.07239	0.03168	134	0.61	031.42	0
031	Phosphorus (%)	2053	1.7850	0.03000	1.7409	0.07239	0.03168	134	0.61	031.02	0
031	Phosphorus (%)	0208	1.7855	0.04100	1.7409	0.07239	0.03168	134	0.62	031.01	0
031	Phosphorus (%)	0893	1.7900	0.00000	1.7409	0.07239	0.03168	134	0.68	031.00	0
031	Phosphorus (%)	0016	1.7900	0.00000	1.7409	0.07239	0.03168	134	0.68	031.01	0
031	Phosphorus (%)	0205	1.7900	0.02000	1.7409	0.07239	0.03168	134	0.68	031.42	0
031	Phosphorus (%)	2004	1.7900	0.02000	1.7409	0.07239	0.03168	134	0.68	031.44	0
031	Phosphorus (%)	0036	1.7939	0.00640	1.7409	0.07239	0.03168	134	0.73	031.03	0
031	Phosphorus (%)	0226	1.7950	0.09000	1.7409	0.07239	0.03168	134	0.75	031.41	0
031	Phosphorus (%)	0510	1.7950	0.01000	1.7409	0.07239	0.03168	134	0.75	031.43	0
031	Phosphorus (%)	0027	1.7975	0.00900	1.7409	0.07239	0.03168	134	0.78	031.43	0
031	Phosphorus (%)	0229	1.8000	0.00800	1.7409	0.07239	0.03168	134	0.82	031.41	0
031	Phosphorus (%)	0848	1.8000	0.00000	1.7409	0.07239	0.03168	134	0.82	031.41	0
031	Phosphorus (%)	0505	1.8050	0.03000	1.7409	0.07239	0.03168	134	0.89	031.02	0
031	Phosphorus (%)	0413	1.8100	0.06000	1.7409	0.07239	0.03168	134	0.95	031.53	0
031	Phosphorus (%)	0553	1.8100	0.08000	1.7409	0.07239	0.03168	134	0.95	031.53	0
031	Phosphorus (%)	0242	1.8100	0.02000	1.7409	0.07239	0.03168	134	0.95	031.99	0
031	Phosphorus (%)	0036	1.8173	0.00120	1.7409	0.07239	0.03168	134	1.06	031.43	0
031	Phosphorus (%)	0098	1.8230	0.10000	1.7409	0.07239	0.03168	134	1.13	031.43	0
031	Phosphorus (%)	0265	1.8250	0.01000	1.7409	0.07239	0.03168	134	1.16	031.42	0
031	Phosphorus (%)	0035	1.8400	0.00000	1.7409	0.07239	0.03168	134	1.37	031.43	0
031	Phosphorus (%)	0964	1.8445	0.10940	1.7409	0.07239	0.03168	134	1.43	031.43	0
031	Phosphorus (%)	2033	1.8550	0.13000	1.7409	0.07239	0.03168	134	1.58	031.43	0
031	Phosphorus (%)	0512	1.8565	0.00900	1.7409	0.07239	0.03168	134	1.60	031.43	0
031	Phosphorus (%)	0425	1.8900	0.06000	1.7409	0.07239	0.03168	134	2.06	031.43	0
031	Phosphorus (%)	0692	1.8950	0.03000	1.7409	0.07239	0.03168	134	2.13	031.42	0
031	Phosphorus (%)	0726	1.9145	0.00900	1.7409	0.07239	0.03168	134	2.40	031.42	0
031	Phosphorus (%)	0190	1.9300	0.02000	1.7409	0.07239	0.03168	134	2.61	031.42	0
031	Phosphorus (%)	0353	1.9500	0.02000	1.7409	0.07239	0.03168	134	2.89	031.43	0
031	Phosphorus (%)	0598	1.9570	0.00600	1.7409	0.07239	0.03168	134	2.99	031.41	0
031	Phosphorus (%)	0572	1.9950	0.03000	1.7409	0.07239	0.03168	134	3.51	031.53	0
031	Phosphorus (%)	0300	2.0200	0.14000	1.7409	0.07239	0.03168	134	3.86	031.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	0358	2.0500	0.08000	1.7409	0.07239	0.03168	134	4.27	031.41	0
031	Phosphorus (%)	0083	2.1050	0.03000	1.7409	0.07239	0.03168	134	5.03	031.43	0
031	Phosphorus (%)	0910	1.7200	0.26000	1.7409	0.07239	0.03168	134	-0.29	031.41	1
031	Phosphorus (%)	0687	1.8000	0.20000	1.7409	0.07239	0.03168	134	0.82	031.01	1
031	Phosphorus (%)	0108	0.69500	0.01000	1.7409	0.07239	0.03168	134	-14.45	031.01	2
032	Potassium (%)	0536	0.22000	0.02000	0.57309	0.03434	0.01565	93	-10.28	032.31	0
032	Potassium (%)	0511	0.49500	0.01000	0.57309	0.03434	0.01565	93	-2.27	032.41	0
032	Potassium (%)	0144	0.50000	0.04000	0.57309	0.03434	0.01565	93	-2.13	032.41	0
032	Potassium (%)	2012	0.51000	0.04000	0.57309	0.03434	0.01565	93	-1.84	032.41	0
032	Potassium (%)	0720	0.52000	0.00000	0.57309	0.03434	0.01565	93	-1.55	032.41	0
032	Potassium (%)	0169	0.52500	0.03000	0.57309	0.03434	0.01565	93	-1.40	032.32	0
032	Potassium (%)	2129	0.52550	0.03300	0.57309	0.03434	0.01565	93	-1.39	032.42	0
032	Potassium (%)	0918	0.53400	0.03000	0.57309	0.03434	0.01565	93	-1.14	032.43	0
032	Potassium (%)	0042	0.53600	0.01000	0.57309	0.03434	0.01565	93	-1.08	032.42	0
032	Potassium (%)	0629	0.53750	0.00500	0.57309	0.03434	0.01565	93	-1.04	032.31	0
032	Potassium (%)	0229	0.53950	0.02500	0.57309	0.03434	0.01565	93	-0.98	032.41	0
032	Potassium (%)	0089	0.54000	0.00000	0.57309	0.03434	0.01565	93	-0.96	032.43	0
032	Potassium (%)	0353	0.54000	0.02000	0.57309	0.03434	0.01565	93	-0.96	032.43	0
032	Potassium (%)	0675	0.54000	0.02000	0.57309	0.03434	0.01565	93	-0.96	032.43	0
032	Potassium (%)	0682	0.54000	0.00000	0.57309	0.03434	0.01565	93	-0.96	032.43	0
032	Potassium (%)	0968	0.54400	0.00400	0.57309	0.03434	0.01565	93	-0.85	032.43	0
032	Potassium (%)	0035	0.54500	0.05000	0.57309	0.03434	0.01565	93	-0.82	032.42	0
032	Potassium (%)	0035	0.54500	0.05000	0.57309	0.03434	0.01565	93	-0.82	032.43	0
032	Potassium (%)	2022	0.54500	0.00200	0.57309	0.03434	0.01565	93	-0.82	032.31	0
032	Potassium (%)	0045	0.54500	0.02800	0.57309	0.03434	0.01565	93	-0.82	032.42	0
032	Potassium (%)	0693	0.54800	0.01600	0.57309	0.03434	0.01565	93	-0.73	032.42	0
032	Potassium (%)	0148	0.54950	0.01700	0.57309	0.03434	0.01565	93	-0.69	032.41	0
032	Potassium (%)	0034	0.54950	0.00500	0.57309	0.03434	0.01565	93	-0.69	032.43	0
032	Potassium (%)	0553	0.54950	0.01700	0.57309	0.03434	0.01565	93	-0.69	032.53	0
032	Potassium (%)	0019	0.55000	0.02000	0.57309	0.03434	0.01565	93	-0.67	032.41	0
032	Potassium (%)	0948	0.55050	0.00700	0.57309	0.03434	0.01565	93	-0.66	032.31	0
032	Potassium (%)	0955	0.55150	0.01700	0.57309	0.03434	0.01565	93	-0.63	032.44	0
032	Potassium (%)	0154	0.55250	0.01960	0.57309	0.03434	0.01565	93	-0.60	032.52	0
032	Potassium (%)	0861	0.55350	0.00300	0.57309	0.03434	0.01565	93	-0.57	032.43	0
032	Potassium (%)	0171	0.55400	0.00600	0.57309	0.03434	0.01565	93	-0.56	032.41	0
032	Potassium (%)	0014	0.55450	0.05100	0.57309	0.03434	0.01565	93	-0.54	032.31	0
032	Potassium (%)	0345	0.55500	0.01000	0.57309	0.03434	0.01565	93	-0.53	032.43	0
032	Potassium (%)	0208	0.55650	0.00500	0.57309	0.03434	0.01565	93	-0.48	032.41	0
032	Potassium (%)	0011	0.55665	0.00450	0.57309	0.03434	0.01565	93	-0.48	032.41	0
032	Potassium (%)	0354	0.56000	0.00000	0.57309	0.03434	0.01565	93	-0.38	032.31	0
032	Potassium (%)	0910	0.56000	0.06000	0.57309	0.03434	0.01565	93	-0.38	032.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
032	Potassium (%)	0510	0.56000	0.00000	0.57309	0.03434	0.01565	93	-0.38	032.43	0
032	Potassium (%)	0650	0.56200	0.01600	0.57309	0.03434	0.01565	93	-0.32	032.31	0
032	Potassium (%)	0186	0.56200	0.02400	0.57309	0.03434	0.01565	93	-0.32	032.52	0
032	Potassium (%)	0619	0.56250	0.00700	0.57309	0.03434	0.01565	93	-0.31	032.41	0
032	Potassium (%)	0294	0.56500	0.03000	0.57309	0.03434	0.01565	93	-0.24	032.42	0
032	Potassium (%)	0425	0.56500	0.03000	0.57309	0.03434	0.01565	93	-0.24	032.43	0
032	Potassium (%)	0100	0.56500	0.01000	0.57309	0.03434	0.01565	93	-0.24	032.99	0
032	Potassium (%)	0874	0.56550	0.03700	0.57309	0.03434	0.01565	93	-0.22	032.31	0
032	Potassium (%)	0142	0.56725	0.00630	0.57309	0.03434	0.01565	93	-0.17	032.31	0
032	Potassium (%)	0164	0.56750	0.00300	0.57309	0.03434	0.01565	93	-0.16	032.41	0
032	Potassium (%)	0175	0.57000	0.02000	0.57309	0.03434	0.01565	93	-0.09	032.31	0
032	Potassium (%)	0051	0.57110	0.01880	0.57309	0.03434	0.01565	93	-0.06	032.41	0
032	Potassium (%)	0205	0.57250	0.00300	0.57309	0.03434	0.01565	93	-0.02	032.42	0
032	Potassium (%)	0043	0.57300	0.00800	0.57309	0.03434	0.01565	93	0.00	032.43	0
032	Potassium (%)	0598	0.57465	0.00690	0.57309	0.03434	0.01565	93	0.05	032.41	0
032	Potassium (%)	0520	0.57500	0.01000	0.57309	0.03434	0.01565	93	0.06	032.41	0
032	Potassium (%)	0560	0.57550	0.00200	0.57309	0.03434	0.01565	93	0.07	032.42	0
032	Potassium (%)	0870	0.57940	0.00420	0.57309	0.03434	0.01565	93	0.18	032.42	0
032	Potassium (%)	0878	0.58000	0.00000	0.57309	0.03434	0.01565	93	0.20	032.31	0
032	Potassium (%)	0027	0.58000	0.00400	0.57309	0.03434	0.01565	93	0.20	032.43	0
032	Potassium (%)	0208	0.58050	0.00700	0.57309	0.03434	0.01565	93	0.22	032.31	0
032	Potassium (%)	0037	0.58500	0.01000	0.57309	0.03434	0.01565	93	0.35	032.42	0
032	Potassium (%)	0265	0.58500	0.03000	0.57309	0.03434	0.01565	93	0.35	032.42	0
032	Potassium (%)	0278	0.58500	0.03000	0.57309	0.03434	0.01565	93	0.35	032.42	0
032	Potassium (%)	0083	0.58500	0.01000	0.57309	0.03434	0.01565	93	0.35	032.43	0
032	Potassium (%)	0242	0.58500	0.01000	0.57309	0.03434	0.01565	93	0.35	032.99	0
032	Potassium (%)	0512	0.58750	0.00120	0.57309	0.03434	0.01565	93	0.42	032.43	0
032	Potassium (%)	0508	0.58780	0.01960	0.57309	0.03434	0.01565	93	0.43	032.43	0
032	Potassium (%)	0049	0.59000	0.00000	0.57309	0.03434	0.01565	93	0.49	032.41	0
032	Potassium (%)	0098	0.59000	0.04000	0.57309	0.03434	0.01565	93	0.49	032.43	0
032	Potassium (%)	2033	0.59000	0.00000	0.57309	0.03434	0.01565	93	0.49	032.43	0
032	Potassium (%)	0139	0.59250	0.02500	0.57309	0.03434	0.01565	93	0.57	032.41	0
032	Potassium (%)	0006	0.59250	0.00500	0.57309	0.03434	0.01565	93	0.57	032.43	0
032	Potassium (%)	0186	0.59400	0.00400	0.57309	0.03434	0.01565	93	0.61	032.42	0
032	Potassium (%)	2053	0.59500	0.01000	0.57309	0.03434	0.01565	93	0.64	032.31	0
032	Potassium (%)	0226	0.59500	0.01000	0.57309	0.03434	0.01565	93	0.64	032.41	0
032	Potassium (%)	0413	0.60000	0.00000	0.57309	0.03434	0.01565	93	0.78	032.41	0
032	Potassium (%)	2089	0.60000	0.00000	0.57309	0.03434	0.01565	93	0.78	032.43	0
032	Potassium (%)	0017	0.60150	0.01900	0.57309	0.03434	0.01565	93	0.83	032.43	0
032	Potassium (%)	0098	0.60500	0.01000	0.57309	0.03434	0.01565	93	0.93	032.41	0
032	Potassium (%)	0964	0.60730	0.07060	0.57309	0.03434	0.01565	93	1.00	032.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
032	Potassium (%)	2004	0.60900	0.01000	0.57309	0.03434	0.01565	93	1.05	032.44	0
032	Potassium (%)	0413	0.61000	0.00000	0.57309	0.03434	0.01565	93	1.07	032.53	0
032	Potassium (%)	0297	0.61500	0.03000	0.57309	0.03434	0.01565	93	1.22	032.43	0
032	Potassium (%)	0407	0.61565	0.00230	0.57309	0.03434	0.01565	93	1.24	032.41	0
032	Potassium (%)	0357	0.61925	0.00610	0.57309	0.03434	0.01565	93	1.34	032.42	0
032	Potassium (%)	0692	0.62000	0.02000	0.57309	0.03434	0.01565	93	1.37	032.42	0
032	Potassium (%)	0263	0.62015	0.01710	0.57309	0.03434	0.01565	93	1.37	032.42	0
032	Potassium (%)	0939	0.62500	0.01000	0.57309	0.03434	0.01565	93	1.51	032.31	0
032	Potassium (%)	0358	0.62500	0.01000	0.57309	0.03434	0.01565	93	1.51	032.41	0
032	Potassium (%)	0190	0.62500	0.01000	0.57309	0.03434	0.01565	93	1.51	032.42	0
032	Potassium (%)	0726	0.62680	0.00420	0.57309	0.03434	0.01565	93	1.56	032.42	0
032	Potassium (%)	0405	0.63500	0.03000	0.57309	0.03434	0.01565	93	1.80	032.41	0
032	Potassium (%)	0202	0.63930	0.00260	0.57309	0.03434	0.01565	93	1.93	032.43	0
032	Potassium (%)	0572	0.67900	0.04600	0.57309	0.03434	0.01565	93	3.08	032.53	0
032	Potassium (%)	0366	0.69000	0.02000	0.57309	0.03434	0.01565	93	3.40	032.42	0
032	Potassium (%)	0529	0.72500	0.01000	0.57309	0.03434	0.01565	93	4.42	032.31	0
032	Potassium (%)	0003	0.57000	0.08000	0.57309	0.03434	0.01565	93	-0.09	032.41	1
032	Potassium (%)	0889	0.11025	0.05650	0.57309	0.03434	0.01565	93	-13.48	032.99	2
033	Salt as chloride (%)	0017	0.91700	0.00800	1.2477	0.07276	0.01916	55	-4.55	033.99	0
033	Salt as chloride (%)	0358	0.94500	0.01000	1.2477	0.07276	0.01916	55	-4.16	033.99	0
033	Salt as chloride (%)	0171	1.0700	0.00000	1.2477	0.07276	0.01916	55	-2.44	033.00	0
033	Salt as chloride (%)	0619	1.0800	0.02000	1.2477	0.07276	0.01916	55	-2.31	033.01	0
033	Salt as chloride (%)	2033	1.0850	0.03000	1.2477	0.07276	0.01916	55	-2.24	033.99	0
033	Salt as chloride (%)	0893	1.0900	0.00000	1.2477	0.07276	0.01916	55	-2.17	033.00	0
033	Salt as chloride (%)	0353	1.0950	0.03000	1.2477	0.07276	0.01916	55	-2.10	033.00	0
033	Salt as chloride (%)	0539	1.1350	0.01000	1.2477	0.07276	0.01916	55	-1.55	033.00	0
033	Salt as chloride (%)	0897	1.1450	0.01000	1.2477	0.07276	0.01916	55	-1.41	033.00	0
033	Salt as chloride (%)	0878	1.1600	0.02000	1.2477	0.07276	0.01916	55	-1.21	033.05	0
033	Salt as chloride (%)	0693	1.1630	0.04600	1.2477	0.07276	0.01916	55	-1.16	033.00	0
033	Salt as chloride (%)	0536	1.1700	0.02000	1.2477	0.07276	0.01916	55	-1.07	033.99	0
033	Salt as chloride (%)	0903	1.2050	0.01000	1.2477	0.07276	0.01916	55	-0.59	033.00	0
033	Salt as chloride (%)	0910	1.2050	0.01000	1.2477	0.07276	0.01916	55	-0.59	033.99	0
033	Salt as chloride (%)	0552	1.2100	0.02000	1.2477	0.07276	0.01916	55	-0.52	033.99	0
033	Salt as chloride (%)	0045	1.2150	0.03000	1.2477	0.07276	0.01916	55	-0.45	033.00	0
033	Salt as chloride (%)	0298	1.2200	0.02000	1.2477	0.07276	0.01916	55	-0.38	033.00	0
033	Salt as chloride (%)	2022	1.2310	0.06000	1.2477	0.07276	0.01916	55	-0.23	033.00	0
033	Salt as chloride (%)	2009	1.2323	0.00440	1.2477	0.07276	0.01916	55	-0.21	033.00	0
033	Salt as chloride (%)	0885	1.2342	0.00980	1.2477	0.07276	0.01916	55	-0.19	033.00	0
033	Salt as chloride (%)	0884	1.2400	0.04000	1.2477	0.07276	0.01916	55	-0.11	033.00	0
033	Salt as chloride (%)	0148	1.2445	0.00100	1.2477	0.07276	0.01916	55	-0.04	033.01	0
033	Salt as chloride (%)	2129	1.2495	0.03100	1.2477	0.07276	0.01916	55	0.02	033.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
033	Salt as chloride (%)	0689	1.2500	0.04000	1.2477	0.07276	0.01916	55	0.03	033.05	0
033	Salt as chloride (%)	0682	1.2500	0.00000	1.2477	0.07276	0.01916	55	0.03	033.99	0
033	Salt as chloride (%)	0675	1.2600	0.04000	1.2477	0.07276	0.01916	55	0.17	033.00	0
033	Salt as chloride (%)	0366	1.2650	0.03000	1.2477	0.07276	0.01916	55	0.24	033.00	0
033	Salt as chloride (%)	0590	1.2650	0.01000	1.2477	0.07276	0.01916	55	0.24	033.01	0
033	Salt as chloride (%)	0083	1.2700	0.02000	1.2477	0.07276	0.01916	55	0.31	033.01	0
033	Salt as chloride (%)	0309	1.2720	0.00170	1.2477	0.07276	0.01916	55	0.33	033.00	0
033	Salt as chloride (%)	0650	1.2750	0.07000	1.2477	0.07276	0.01916	55	0.37	033.01	0
033	Salt as chloride (%)	0683	1.2750	0.03000	1.2477	0.07276	0.01916	55	0.37	033.01	0
033	Salt as chloride (%)	2062	1.2800	0.02000	1.2477	0.07276	0.01916	55	0.44	033.00	0
033	Salt as chloride (%)	0226	1.2800	0.02000	1.2477	0.07276	0.01916	55	0.44	033.01	0
033	Salt as chloride (%)	0242	1.2800	0.02000	1.2477	0.07276	0.01916	55	0.44	033.01	0
033	Salt as chloride (%)	0407	1.2800	0.00400	1.2477	0.07276	0.01916	55	0.44	033.01	0
033	Salt as chloride (%)	0425	1.2800	0.00000	1.2477	0.07276	0.01916	55	0.44	033.01	0
033	Salt as chloride (%)	0510	1.2800	0.02000	1.2477	0.07276	0.01916	55	0.44	033.01	0
033	Salt as chloride (%)	0354	1.2900	0.00000	1.2477	0.07276	0.01916	55	0.58	033.01	0
033	Salt as chloride (%)	0948	1.2900	0.00000	1.2477	0.07276	0.01916	55	0.58	033.01	0
033	Salt as chloride (%)	0229	1.2920	0.00400	1.2477	0.07276	0.01916	55	0.61	033.01	0
033	Salt as chloride (%)	0861	1.2950	0.01000	1.2477	0.07276	0.01916	55	0.65	033.01	0
033	Salt as chloride (%)	0205	1.3000	0.02000	1.2477	0.07276	0.01916	55	0.72	033.01	0
033	Salt as chloride (%)	0186	1.3100	0.06000	1.2477	0.07276	0.01916	55	0.86	033.01	0
033	Salt as chloride (%)	0278	1.3100	0.02000	1.2477	0.07276	0.01916	55	0.86	033.01	0
033	Salt as chloride (%)	0629	1.3100	0.00000	1.2477	0.07276	0.01916	55	0.86	033.01	0
033	Salt as chloride (%)	0644	1.3118	0.00390	1.2477	0.07276	0.01916	55	0.88	033.01	0
033	Salt as chloride (%)	0895	1.3150	0.01000	1.2477	0.07276	0.01916	55	0.92	033.00	0
033	Salt as chloride (%)	0098	1.3150	0.03000	1.2477	0.07276	0.01916	55	0.92	033.01	0
033	Salt as chloride (%)	0164	1.3200	0.02000	1.2477	0.07276	0.01916	55	0.99	033.01	0
033	Salt as chloride (%)	0175	1.3200	0.00000	1.2477	0.07276	0.01916	55	0.99	033.01	0
033	Salt as chloride (%)	0265	1.3300	0.00000	1.2477	0.07276	0.01916	55	1.13	033.03	0
033	Salt as chloride (%)	0874	1.3360	0.00000	1.2477	0.07276	0.01916	55	1.21	033.01	0
033	Salt as chloride (%)	0413	1.3400	0.00000	1.2477	0.07276	0.01916	55	1.27	033.01	0
033	Salt as chloride (%)	0144	1.3450	0.11000	1.2477	0.07276	0.01916	55	1.34	033.03	0
033	Salt as chloride (%)	2073	1.1500	0.16000	1.2477	0.07276	0.01916	55	-1.34	033.99	1
033	Salt as chloride (%)	0100	2.2300	0.16000	1.2477	0.07276	0.01916	55	13.50	033.01	1
033	Salt as chloride (%)	0169	0.00000	0.00000	1.2477	0.07276	0.01916	55		033.00	4
034	Selenium (mg / kg (ppm))	0692	0.10000	0.00000	1.6467	0.41211	0.10427	17	-3.75	034.42	0
034	Selenium (mg / kg (ppm))	0563	0.25545	0.06670	1.6467	0.41211	0.10427	17	-3.38	034.04	0
034	Selenium (mg / kg (ppm))	0964	1.3048	0.28710	1.6467	0.41211	0.10427	17	-0.83	034.43	0
034	Selenium (mg / kg (ppm))	0045	1.3400	0.04000	1.6467	0.41211	0.10427	17	-0.74	034.04	0
034	Selenium (mg / kg (ppm))	0171	1.4550	0.09000	1.6467	0.41211	0.10427	17	-0.47	034.04	0
034	Selenium (mg / kg (ppm))	0553	1.5500	0.02000	1.6467	0.41211	0.10427	17	-0.23	034.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
034	Selenium (mg / kg (ppm))	0560	1.5600	0.04000	1.6467	0.41211	0.10427	17	-0.21	034.53	0
034	Selenium (mg / kg (ppm))	0164	1.5750	0.35000	1.6467	0.41211	0.10427	17	-0.17	034.04	0
034	Selenium (mg / kg (ppm))	0169	1.6200	0.00000	1.6467	0.41211	0.10427	17	-0.06	034.04	0
034	Selenium (mg / kg (ppm))	0038	1.6500	0.08000	1.6467	0.41211	0.10427	17	0.01	034.01	0
034	Selenium (mg / kg (ppm))	2012	1.7000	0.02000	1.6467	0.41211	0.10427	17	0.13	034.53	0
034	Selenium (mg / kg (ppm))	0098	1.7520	0.02400	1.6467	0.41211	0.10427	17	0.26	034.53	0
034	Selenium (mg / kg (ppm))	0910	1.8350	0.01000	1.6467	0.41211	0.10427	17	0.46	034.52	0
034	Selenium (mg / kg (ppm))	2033	2.0700	0.06000	1.6467	0.41211	0.10427	17	1.03	034.43	0
034	Selenium (mg / kg (ppm))	0918	2.1255	0.08480	1.6467	0.41211	0.10427	17	1.16	034.53	0
034	Selenium (mg / kg (ppm))	0413	2.1350	0.17000	1.6467	0.41211	0.10427	17	1.18	034.53	0
034	Selenium (mg / kg (ppm))	0186	2.3150	0.43000	1.6467	0.41211	0.10427	17	1.62	034.52	0
035	Sodium (%)	0889	0.20255	0.02830	0.34557	0.02112	0.00945	98	-6.77	035.99	0
035	Sodium (%)	0405	0.27300	0.00600	0.34557	0.02112	0.00945	98	-3.44	035.41	0
035	Sodium (%)	0034	0.28565	0.00030	0.34557	0.02112	0.00945	98	-2.84	035.43	0
035	Sodium (%)	0300	0.30000	0.02000	0.34557	0.02112	0.00945	98	-2.16	035.33	0
035	Sodium (%)	0878	0.30100	0.00200	0.34557	0.02112	0.00945	98	-2.11	035.01	0
035	Sodium (%)	0043	0.30450	0.00500	0.34557	0.02112	0.00945	98	-1.94	035.43	0
035	Sodium (%)	0536	0.30500	0.01000	0.34557	0.02112	0.00945	98	-1.92	035.31	0
035	Sodium (%)	0169	0.30500	0.00200	0.34557	0.02112	0.00945	98	-1.92	035.32	0
035	Sodium (%)	2006	0.30900	0.00000	0.34557	0.02112	0.00945	98	-1.73	035.01	0
035	Sodium (%)	0045	0.31100	0.03000	0.34557	0.02112	0.00945	98	-1.64	035.42	0
035	Sodium (%)	0148	0.31650	0.01900	0.34557	0.02112	0.00945	98	-1.38	035.41	0
035	Sodium (%)	0510	0.31900	0.00400	0.34557	0.02112	0.00945	98	-1.26	035.43	0
035	Sodium (%)	0720	0.32000	0.00000	0.34557	0.02112	0.00945	98	-1.21	035.41	0
035	Sodium (%)	0650	0.32100	0.02000	0.34557	0.02112	0.00945	98	-1.16	035.31	0
035	Sodium (%)	0345	0.32250	0.00300	0.34557	0.02112	0.00945	98	-1.09	035.43	0
035	Sodium (%)	0689	0.32500	0.03000	0.34557	0.02112	0.00945	98	-0.97	035.31	0
035	Sodium (%)	2012	0.32500	0.01000	0.34557	0.02112	0.00945	98	-0.97	035.41	0
035	Sodium (%)	0035	0.32500	0.01000	0.34557	0.02112	0.00945	98	-0.97	035.43	0
035	Sodium (%)	0693	0.32700	0.00400	0.34557	0.02112	0.00945	98	-0.88	035.42	0
035	Sodium (%)	0563	0.32725	0.00250	0.34557	0.02112	0.00945	98	-0.87	035.31	0
035	Sodium (%)	2022	0.32800	0.00400	0.34557	0.02112	0.00945	98	-0.83	035.31	0
035	Sodium (%)	0918	0.32950	0.02100	0.34557	0.02112	0.00945	98	-0.76	035.43	0
035	Sodium (%)	0144	0.33000	0.04000	0.34557	0.02112	0.00945	98	-0.74	035.41	0
035	Sodium (%)	0910	0.33000	0.04000	0.34557	0.02112	0.00945	98	-0.74	035.41	0
035	Sodium (%)	0861	0.33050	0.00900	0.34557	0.02112	0.00945	98	-0.71	035.43	0
035	Sodium (%)	0205	0.33200	0.00000	0.34557	0.02112	0.00945	98	-0.64	035.42	0
035	Sodium (%)	0123	0.33500	0.01000	0.34557	0.02112	0.00945	98	-0.50	035.41	0
035	Sodium (%)	0035	0.33500	0.01000	0.34557	0.02112	0.00945	98	-0.50	035.42	0
035	Sodium (%)	0164	0.33650	0.00300	0.34557	0.02112	0.00945	98	-0.43	035.41	0
035	Sodium (%)	0629	0.33750	0.00500	0.34557	0.02112	0.00945	98	-0.38	035.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
035	Sodium (%)	0520	0.33760	0.00580	0.34557	0.02112	0.00945	98	-0.38	035.41	0
035	Sodium (%)	0011	0.33880	0.00220	0.34557	0.02112	0.00945	98	-0.32	035.41	0
035	Sodium (%)	0298	0.34000	0.02000	0.34557	0.02112	0.00945	98	-0.26	035.41	0
035	Sodium (%)	0098	0.34000	0.02000	0.34557	0.02112	0.00945	98	-0.26	035.43	0
035	Sodium (%)	0278	0.34000	0.00000	0.34557	0.02112	0.00945	98	-0.26	035.42	0
035	Sodium (%)	0682	0.34000	0.00000	0.34557	0.02112	0.00945	98	-0.26	035.43	0
035	Sodium (%)	0051	0.34015	0.00910	0.34557	0.02112	0.00945	98	-0.26	035.41	0
035	Sodium (%)	0208	0.34100	0.00200	0.34557	0.02112	0.00945	98	-0.22	035.41	0
035	Sodium (%)	0154	0.34145	0.00770	0.34557	0.02112	0.00945	98	-0.20	035.52	0
035	Sodium (%)	0142	0.34190	0.00400	0.34557	0.02112	0.00945	98	-0.17	035.31	0
035	Sodium (%)	0560	0.34375	0.00410	0.34557	0.02112	0.00945	98	-0.09	035.42	0
035	Sodium (%)	0175	0.34500	0.01000	0.34557	0.02112	0.00945	98	-0.03	035.31	0
035	Sodium (%)	0233	0.34500	0.01000	0.34557	0.02112	0.00945	98	-0.03	035.31	0
035	Sodium (%)	0878	0.34500	0.01000	0.34557	0.02112	0.00945	98	-0.03	035.31	0
035	Sodium (%)	0939	0.34500	0.01000	0.34557	0.02112	0.00945	98	-0.03	035.31	0
035	Sodium (%)	0019	0.34500	0.01000	0.34557	0.02112	0.00945	98	-0.03	035.41	0
035	Sodium (%)	0098	0.34500	0.01000	0.34557	0.02112	0.00945	98	-0.03	035.41	0
035	Sodium (%)	0171	0.34500	0.01000	0.34557	0.02112	0.00945	98	-0.03	035.41	0
035	Sodium (%)	0413	0.34500	0.01000	0.34557	0.02112	0.00945	98	-0.03	035.41	0
035	Sodium (%)	0265	0.34500	0.01000	0.34557	0.02112	0.00945	98	-0.03	035.42	0
035	Sodium (%)	0294	0.34500	0.01000	0.34557	0.02112	0.00945	98	-0.03	035.42	0
035	Sodium (%)	0242	0.34500	0.03000	0.34557	0.02112	0.00945	98	-0.03	035.99	0
035	Sodium (%)	0512	0.34505	0.00070	0.34557	0.02112	0.00945	98	-0.02	035.43	0
035	Sodium (%)	0229	0.34800	0.01000	0.34557	0.02112	0.00945	98	0.12	035.41	0
035	Sodium (%)	0955	0.34950	0.01900	0.34557	0.02112	0.00945	98	0.19	035.41	0
035	Sodium (%)	0354	0.35000	0.00000	0.34557	0.02112	0.00945	98	0.21	035.31	0
035	Sodium (%)	0366	0.35000	0.00000	0.34557	0.02112	0.00945	98	0.21	035.42	0
035	Sodium (%)	0874	0.35100	0.01000	0.34557	0.02112	0.00945	98	0.26	035.31	0
035	Sodium (%)	0407	0.35135	0.00390	0.34557	0.02112	0.00945	98	0.27	035.41	0
035	Sodium (%)	0226	0.35150	0.00900	0.34557	0.02112	0.00945	98	0.28	035.41	0
035	Sodium (%)	0186	0.35200	0.00000	0.34557	0.02112	0.00945	98	0.30	035.42	0
035	Sodium (%)	0006	0.35400	0.00400	0.34557	0.02112	0.00945	98	0.40	035.43	0
035	Sodium (%)	0027	0.35450	0.00500	0.34557	0.02112	0.00945	98	0.42	035.43	0
035	Sodium (%)	0870	0.35455	0.00670	0.34557	0.02112	0.00945	98	0.43	035.42	0
035	Sodium (%)	0108	0.35500	0.01000	0.34557	0.02112	0.00945	98	0.45	035.05	0
035	Sodium (%)	0003	0.35500	0.01000	0.34557	0.02112	0.00945	98	0.45	035.41	0
035	Sodium (%)	0049	0.35500	0.01000	0.34557	0.02112	0.00945	98	0.45	035.41	0
035	Sodium (%)	0083	0.35500	0.01000	0.34557	0.02112	0.00945	98	0.45	035.43	0
035	Sodium (%)	0297	0.35500	0.01000	0.34557	0.02112	0.00945	98	0.45	035.43	0
035	Sodium (%)	0675	0.35500	0.03000	0.34557	0.02112	0.00945	98	0.45	035.43	0
035	Sodium (%)	0208	0.35550	0.01500	0.34557	0.02112	0.00945	98	0.47	035.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
035	Sodium (%)	0529	0.35550	0.00100	0.34557	0.02112	0.00945	98	0.47	035.31	0
035	Sodium (%)	0572	0.35600	0.00800	0.34557	0.02112	0.00945	98	0.49	035.53	0
035	Sodium (%)	0202	0.35980	0.00960	0.34557	0.02112	0.00945	98	0.67	035.43	0
035	Sodium (%)	0884	0.36000	0.00000	0.34557	0.02112	0.00945	98	0.68	035.02	0
035	Sodium (%)	0353	0.36000	0.00000	0.34557	0.02112	0.00945	98	0.68	035.43	0
035	Sodium (%)	0413	0.36000	0.02000	0.34557	0.02112	0.00945	98	0.68	035.53	0
035	Sodium (%)	0100	0.36000	0.00000	0.34557	0.02112	0.00945	98	0.68	035.99	0
035	Sodium (%)	0017	0.36050	0.00300	0.34557	0.02112	0.00945	98	0.71	035.43	0
035	Sodium (%)	2004	0.36100	0.00400	0.34557	0.02112	0.00945	98	0.73	035.41	0
035	Sodium (%)	0692	0.36500	0.01000	0.34557	0.02112	0.00945	98	0.92	035.42	0
035	Sodium (%)	0425	0.36500	0.01000	0.34557	0.02112	0.00945	98	0.92	035.43	0
035	Sodium (%)	0553	0.36700	0.04600	0.34557	0.02112	0.00945	98	1.01	035.53	0
035	Sodium (%)	0964	0.36770	0.03620	0.34557	0.02112	0.00945	98	1.05	035.43	0
035	Sodium (%)	0186	0.36900	0.00800	0.34557	0.02112	0.00945	98	1.11	035.52	0
035	Sodium (%)	0358	0.37000	0.00000	0.34557	0.02112	0.00945	98	1.16	035.41	0
035	Sodium (%)	2033	0.37000	0.00000	0.34557	0.02112	0.00945	98	1.16	035.43	0
035	Sodium (%)	0619	0.37300	0.01000	0.34557	0.02112	0.00945	98	1.30	035.41	0
035	Sodium (%)	0598	0.37350	0.00840	0.34557	0.02112	0.00945	98	1.32	035.41	0
035	Sodium (%)	0508	0.37400	0.00320	0.34557	0.02112	0.00945	98	1.35	035.43	0
035	Sodium (%)	2089	0.37500	0.01000	0.34557	0.02112	0.00945	98	1.39	035.41	0
035	Sodium (%)	0190	0.37500	0.01000	0.34557	0.02112	0.00945	98	1.39	035.42	0
035	Sodium (%)	0726	0.37520	0.00060	0.34557	0.02112	0.00945	98	1.40	035.42	0
035	Sodium (%)	0948	0.37700	0.00200	0.34557	0.02112	0.00945	98	1.49	035.31	0
035	Sodium (%)	0590	0.38000	0.00000	0.34557	0.02112	0.00945	98	1.63	035.99	0
035	Sodium (%)	0263	0.38180	0.00500	0.34557	0.02112	0.00945	98	1.72	035.42	0
035	Sodium (%)	0139	0.38250	0.00500	0.34557	0.02112	0.00945	98	1.75	035.41	0
035	Sodium (%)	0089	0.40000	0.00000	0.34557	0.02112	0.00945	98	2.58	035.43	0
035	Sodium (%)	0037	0.27000	0.06000	0.34557	0.02112	0.00945	98	-3.58	035.05	1
035	Sodium (%)	0042	0.35500	0.08200	0.34557	0.02112	0.00945	98	0.45	035.42	1
035	Sodium (%)	2097	0.67380	0.01340	0.34557	0.02112	0.00945	98	15.54	035.31	2
036	Sulfur (%)	0345	0.69450	0.00300	0.75775	0.04224	0.01510	37	-1.50	036.43	0
036	Sulfur (%)	0035	0.70000	0.02000	0.75775	0.04224	0.01510	37	-1.37	036.43	0
036	Sulfur (%)	0035	0.70500	0.01000	0.75775	0.04224	0.01510	37	-1.25	036.42	0
036	Sulfur (%)	0186	0.71000	0.05600	0.75775	0.04224	0.01510	37	-1.13	036.42	0
036	Sulfur (%)	0708	0.71200	0.01800	0.75775	0.04224	0.01510	37	-1.08	036.42	0
036	Sulfur (%)	0045	0.71250	0.02100	0.75775	0.04224	0.01510	37	-1.07	036.42	0
036	Sulfur (%)	0693	0.71350	0.00100	0.75775	0.04224	0.01510	37	-1.05	036.42	0
036	Sulfur (%)	0366	0.71500	0.01000	0.75775	0.04224	0.01510	37	-1.01	036.42	0
036	Sulfur (%)	0910	0.72280	0.06740	0.75775	0.04224	0.01510	37	-0.83	036.52	0
036	Sulfur (%)	0918	0.73760	0.02240	0.75775	0.04224	0.01510	37	-0.48	036.43	0
036	Sulfur (%)	0870	0.73855	0.00170	0.75775	0.04224	0.01510	37	-0.45	036.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
036	Sulfur (%)	0692	0.74000	0.02000	0.75775	0.04224	0.01510	37	-0.42	036.42	0
036	Sulfur (%)	0229	0.74150	0.01100	0.75775	0.04224	0.01510	37	-0.38	036.04	0
036	Sulfur (%)	0407	0.74150	0.00320	0.75775	0.04224	0.01510	37	-0.38	036.42	0
036	Sulfur (%)	2089	0.75000	0.00000	0.75775	0.04224	0.01510	37	-0.18	036.43	0
036	Sulfur (%)	0186	0.75200	0.02800	0.75775	0.04224	0.01510	37	-0.14	036.52	0
036	Sulfur (%)	0171	0.75450	0.00300	0.75775	0.04224	0.01510	37	-0.08	036.42	0
036	Sulfur (%)	0265	0.75500	0.01000	0.75775	0.04224	0.01510	37	-0.07	036.42	0
036	Sulfur (%)	0560	0.75865	0.00550	0.75775	0.04224	0.01510	37	0.02	036.42	0
036	Sulfur (%)	0297	0.76000	0.00000	0.75775	0.04224	0.01510	37	0.05	036.43	0
036	Sulfur (%)	0202	0.76185	0.00350	0.75775	0.04224	0.01510	37	0.10	036.43	0
036	Sulfur (%)	0553	0.76300	0.04000	0.75775	0.04224	0.01510	37	0.12	036.53	0
036	Sulfur (%)	0353	0.76500	0.01000	0.75775	0.04224	0.01510	37	0.17	036.43	0
036	Sulfur (%)	0964	0.76640	0.00040	0.75775	0.04224	0.01510	37	0.20	036.43	0
036	Sulfur (%)	0242	0.77000	0.00000	0.75775	0.04224	0.01510	37	0.29	036.99	0
036	Sulfur (%)	0098	0.77480	0.06300	0.75775	0.04224	0.01510	37	0.40	036.43	0
036	Sulfur (%)	2053	0.77500	0.01000	0.75775	0.04224	0.01510	37	0.41	036.04	0
036	Sulfur (%)	0278	0.77500	0.01000	0.75775	0.04224	0.01510	37	0.41	036.42	0
036	Sulfur (%)	0357	0.77640	0.01320	0.75775	0.04224	0.01510	37	0.44	036.42	0
036	Sulfur (%)	0169	0.78000	0.02000	0.75775	0.04224	0.01510	37	0.53	036.43	0
036	Sulfur (%)	0889	0.79525	0.02070	0.75775	0.04224	0.01510	37	0.89	036.99	0
036	Sulfur (%)	0294	0.81500	0.01000	0.75775	0.04224	0.01510	37	1.36	036.42	0
036	Sulfur (%)	2033	0.82000	0.00000	0.75775	0.04224	0.01510	37	1.47	036.43	0
036	Sulfur (%)	0190	0.83000	0.02000	0.75775	0.04224	0.01510	37	1.71	036.42	0
036	Sulfur (%)	0510	0.83500	0.01000	0.75775	0.04224	0.01510	37	1.83	036.43	0
036	Sulfur (%)	0726	0.84625	0.00670	0.75775	0.04224	0.01510	37	2.10	036.42	0
036	Sulfur (%)	0037	0.87500	0.01000	0.75775	0.04224	0.01510	37	2.78	036.42	0
036	Sulfur (%)	0226	0.97000	0.10000	0.75775	0.04224	0.01510	37	5.02	036.04	1
037	Zinc (mg / kg (ppm))	0047	289.25	9.5900	364.41	25.173	8.0669	91	-2.99	037.52	0
037	Zinc (mg / kg (ppm))	0870	315.00	4.0000	364.41	25.173	8.0669	91	-1.96	037.42	0
037	Zinc (mg / kg (ppm))	0848	317.31	2.1300	364.41	25.173	8.0669	91	-1.87	037.41	0
037	Zinc (mg / kg (ppm))	0148	318.10	15.800	364.41	25.173	8.0669	91	-1.84	037.41	0
037	Zinc (mg / kg (ppm))	0263	322.00	0.60000	364.41	25.173	8.0669	91	-1.68	037.42	0
037	Zinc (mg / kg (ppm))	2129	326.00	24.800	364.41	25.173	8.0669	91	-1.53	037.99	0
037	Zinc (mg / kg (ppm))	0011	328.98	31.240	364.41	25.173	8.0669	91	-1.41	037.41	0
037	Zinc (mg / kg (ppm))	0186	329.00	4.0000	364.41	25.173	8.0669	91	-1.41	037.42	0
037	Zinc (mg / kg (ppm))	0939	331.66	1.1400	364.41	25.173	8.0669	91	-1.30	037.31	0
037	Zinc (mg / kg (ppm))	0164	333.50	7.0000	364.41	25.173	8.0669	91	-1.23	037.41	0
037	Zinc (mg / kg (ppm))	0353	334.00	24.000	364.41	25.173	8.0669	91	-1.21	037.43	0
037	Zinc (mg / kg (ppm))	0520	336.00	4.0000	364.41	25.173	8.0669	91	-1.13	037.41	0
037	Zinc (mg / kg (ppm))	0045	336.50	7.0000	364.41	25.173	8.0669	91	-1.11	037.42	0
037	Zinc (mg / kg (ppm))	0345	336.50	7.0000	364.41	25.173	8.0669	91	-1.11	037.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
037	Zinc (mg / kg (ppm))	0175	338.50	13.000	364.41	25.173	8.0669	91	-1.03	037.31	0
037	Zinc (mg / kg (ppm))	2144	341.88	0.56000	364.41	25.173	8.0669	91	-0.90	037.99	0
037	Zinc (mg / kg (ppm))	0034	341.95	12.900	364.41	25.173	8.0669	91	-0.89	037.43	0
037	Zinc (mg / kg (ppm))	0918	343.79	1.2480	364.41	25.173	8.0669	91	-0.82	037.43	0
037	Zinc (mg / kg (ppm))	0955	346.00	2.0000	364.41	25.173	8.0669	91	-0.73	037.44	0
037	Zinc (mg / kg (ppm))	0889	346.73	3.7688	364.41	25.173	8.0669	91	-0.70	037.99	0
037	Zinc (mg / kg (ppm))	0043	347.00	2.0000	364.41	25.173	8.0669	91	-0.69	037.43	0
037	Zinc (mg / kg (ppm))	0051	347.17	4.9900	364.41	25.173	8.0669	91	-0.69	037.41	0
037	Zinc (mg / kg (ppm))	0511	347.50	17.000	364.41	25.173	8.0669	91	-0.67	037.41	0
037	Zinc (mg / kg (ppm))	0278	348.25	3.9000	364.41	25.173	8.0669	91	-0.64	037.42	0
037	Zinc (mg / kg (ppm))	0017	348.65	11.975	364.41	25.173	8.0669	91	-0.63	037.43	0
037	Zinc (mg / kg (ppm))	0208	349.65	23.500	364.41	25.173	8.0669	91	-0.59	037.41	0
037	Zinc (mg / kg (ppm))	0405	350.12	0.19000	364.41	25.173	8.0669	91	-0.57	037.41	0
037	Zinc (mg / kg (ppm))	0693	350.26	7.3750	364.41	25.173	8.0669	91	-0.56	037.42	0
037	Zinc (mg / kg (ppm))	0675	350.42	17.290	364.41	25.173	8.0669	91	-0.56	037.43	0
037	Zinc (mg / kg (ppm))	0689	350.75	0.70000	364.41	25.173	8.0669	91	-0.54	037.31	0
037	Zinc (mg / kg (ppm))	2022	351.00	4.0000	364.41	25.173	8.0669	91	-0.53	037.31	0
037	Zinc (mg / kg (ppm))	0629	354.50	1.0000	364.41	25.173	8.0669	91	-0.39	037.31	0
037	Zinc (mg / kg (ppm))	0049	356.11	6.2600	364.41	25.173	8.0669	91	-0.33	037.41	0
037	Zinc (mg / kg (ppm))	0366	357.00	10.000	364.41	25.173	8.0669	91	-0.29	037.42	0
037	Zinc (mg / kg (ppm))	0413	357.00	2.0000	364.41	25.173	8.0669	91	-0.29	037.53	0
037	Zinc (mg / kg (ppm))	0242	358.50	9.0000	364.41	25.173	8.0669	91	-0.23	037.99	0
037	Zinc (mg / kg (ppm))	0563	358.53	2.6355	364.41	25.173	8.0669	91	-0.23	037.31	0
037	Zinc (mg / kg (ppm))	0619	359.00	8.0000	364.41	25.173	8.0669	91	-0.22	037.41	0
037	Zinc (mg / kg (ppm))	0354	360.35	0.34000	364.41	25.173	8.0669	91	-0.16	037.31	0
037	Zinc (mg / kg (ppm))	0720	361.05	3.8200	364.41	25.173	8.0669	91	-0.13	037.41	0
037	Zinc (mg / kg (ppm))	0682	361.62	0.00000	364.41	25.173	8.0669	91	-0.11	037.43	0
037	Zinc (mg / kg (ppm))	2089	362.62	12.820	364.41	25.173	8.0669	91	-0.07	037.43	0
037	Zinc (mg / kg (ppm))	0536	363.05	11.500	364.41	25.173	8.0669	91	-0.05	037.31	0
037	Zinc (mg / kg (ppm))	0510	363.50	7.0000	364.41	25.173	8.0669	91	-0.04	037.43	0
037	Zinc (mg / kg (ppm))	0202	364.20	2.0000	364.41	25.173	8.0669	91	-0.01	037.43	0
037	Zinc (mg / kg (ppm))	0692	367.00	6.0000	364.41	25.173	8.0669	91	0.10	037.42	0
037	Zinc (mg / kg (ppm))	0019	367.40	0.00000	364.41	25.173	8.0669	91	0.12	037.41	0
037	Zinc (mg / kg (ppm))	0968	367.50	5.0000	364.41	25.173	8.0669	91	0.12	037.43	0
037	Zinc (mg / kg (ppm))	0964	367.94	26.830	364.41	25.173	8.0669	91	0.14	037.43	0
037	Zinc (mg / kg (ppm))	2004	368.00	6.0000	364.41	25.173	8.0669	91	0.14	037.44	0
037	Zinc (mg / kg (ppm))	2062	368.05	3.1000	364.41	25.173	8.0669	91	0.14	037.31	0
037	Zinc (mg / kg (ppm))	0171	368.05	2.1000	364.41	25.173	8.0669	91	0.14	037.41	0
037	Zinc (mg / kg (ppm))	0508	369.70	3.0200	364.41	25.173	8.0669	91	0.21	037.43	0
037	Zinc (mg / kg (ppm))	2149	370.76	1.7400	364.41	25.173	8.0669	91	0.25	037.31	0
037	Zinc (mg / kg (ppm))	0560	370.85	31.700	364.41	25.173	8.0669	91	0.26	037.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
037	Zinc (mg / kg (ppm))	0861	371.00	2.0000	364.41	25.173	8.0669	91	0.26	037.43	0
037	Zinc (mg / kg (ppm))	0098	371.55	29.900	364.41	25.173	8.0669	91	0.28	037.43	0
037	Zinc (mg / kg (ppm))	0003	372.00	2.0000	364.41	25.173	8.0669	91	0.30	037.41	0
037	Zinc (mg / kg (ppm))	0294	372.10	9.8000	364.41	25.173	8.0669	91	0.31	037.42	0
037	Zinc (mg / kg (ppm))	0878	372.29	0.11000	364.41	25.173	8.0669	91	0.31	037.31	0
037	Zinc (mg / kg (ppm))	0425	373.60	0.20000	364.41	25.173	8.0669	91	0.36	037.43	0
037	Zinc (mg / kg (ppm))	0037	374.60	1.4000	364.41	25.173	8.0669	91	0.40	037.42	0
037	Zinc (mg / kg (ppm))	2012	374.80	6.0200	364.41	25.173	8.0669	91	0.41	037.41	0
037	Zinc (mg / kg (ppm))	0083	375.50	9.0000	364.41	25.173	8.0669	91	0.44	037.43	0
037	Zinc (mg / kg (ppm))	0186	376.00	10.000	364.41	25.173	8.0669	91	0.46	037.52	0
037	Zinc (mg / kg (ppm))	0265	378.00	18.000	364.41	25.173	8.0669	91	0.54	037.42	0
037	Zinc (mg / kg (ppm))	0413	378.50	15.000	364.41	25.173	8.0669	91	0.56	037.42	0
037	Zinc (mg / kg (ppm))	0098	379.25	5.1000	364.41	25.173	8.0669	91	0.59	037.41	0
037	Zinc (mg / kg (ppm))	0208	381.00	20.000	364.41	25.173	8.0669	91	0.66	037.31	0
037	Zinc (mg / kg (ppm))	0590	381.00	0.00000	364.41	25.173	8.0669	91	0.66	037.99	0
037	Zinc (mg / kg (ppm))	0529	383.20	3.8000	364.41	25.173	8.0669	91	0.75	037.31	0
037	Zinc (mg / kg (ppm))	0407	384.12	0.54410	364.41	25.173	8.0669	91	0.78	037.41	0
037	Zinc (mg / kg (ppm))	0598	384.45	1.1000	364.41	25.173	8.0669	91	0.80	037.41	0
037	Zinc (mg / kg (ppm))	0358	384.75	0.85000	364.41	25.173	8.0669	91	0.81	037.41	0
037	Zinc (mg / kg (ppm))	0910	386.50	27.000	364.41	25.173	8.0669	91	0.88	037.41	0
037	Zinc (mg / kg (ppm))	0229	386.55	0.30000	364.41	25.173	8.0669	91	0.88	037.41	0
037	Zinc (mg / kg (ppm))	0357	387.67	2.5830	364.41	25.173	8.0669	91	0.92	037.42	0
037	Zinc (mg / kg (ppm))	0027	392.04	2.0440	364.41	25.173	8.0669	91	1.10	037.43	0
037	Zinc (mg / kg (ppm))	0874	392.50	15.000	364.41	25.173	8.0669	91	1.12	037.31	0
037	Zinc (mg / kg (ppm))	0226	393.00	18.000	364.41	25.173	8.0669	91	1.14	037.41	0
037	Zinc (mg / kg (ppm))	0297	393.50	11.000	364.41	25.173	8.0669	91	1.16	037.43	0
037	Zinc (mg / kg (ppm))	0512	394.70	4.8000	364.41	25.173	8.0669	91	1.20	037.43	0
037	Zinc (mg / kg (ppm))	0948	403.86	0.43900	364.41	25.173	8.0669	91	1.57	037.33	0
037	Zinc (mg / kg (ppm))	2053	404.00	10.000	364.41	25.173	8.0669	91	1.57	037.42	0
037	Zinc (mg / kg (ppm))	0553	407.50	27.000	364.41	25.173	8.0669	91	1.71	037.53	0
037	Zinc (mg / kg (ppm))	0190	407.92	3.2400	364.41	25.173	8.0669	91	1.73	037.42	0
037	Zinc (mg / kg (ppm))	0572	415.50	9.0000	364.41	25.173	8.0669	91	2.03	037.53	0
037	Zinc (mg / kg (ppm))	0035	423.00	16.000	364.41	25.173	8.0669	91	2.33	037.42	0
037	Zinc (mg / kg (ppm))	0726	423.25	1.3000	364.41	25.173	8.0669	91	2.34	037.42	0
037	Zinc (mg / kg (ppm))	0100	424.50	3.0000	364.41	25.173	8.0669	91	2.39	037.99	0
037	Zinc (mg / kg (ppm))	0169	492.00	18.000	364.41	25.173	8.0669	91	5.07	037.43	0
037	Zinc (mg / kg (ppm))	2033	204.80	340.40	364.41	25.173	8.0669	91	-6.34	037.43	1
037	Zinc (mg / kg (ppm))	0042	465.50	237.00	364.41	25.173	8.0669	91	4.02	037.42	1
038	Molybdenum (mg / kg (ppm))	0353	0.37180	0.23480	0.69522	0.15536	0.05881	18	-2.08	038.43	0
038	Molybdenum (mg / kg (ppm))	0508	0.51210	0.22020	0.69522	0.15536	0.05881	18	-1.18	038.43	0
038	Molybdenum (mg / kg (ppm))	0171	0.55250	0.00500	0.69522	0.15536	0.05881	18	-0.92	038.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
038	Molybdenum (mg / kg (ppm))	0693	0.55900	0.00000	0.69522	0.15536	0.05881	18	-0.88	038.42	0
038	Molybdenum (mg / kg (ppm))	0345	0.56500	0.03000	0.69522	0.15536	0.05881	18	-0.84	038.43	0
038	Molybdenum (mg / kg (ppm))	0169	0.63500	0.01000	0.69522	0.15536	0.05881	18	-0.39	038.43	0
038	Molybdenum (mg / kg (ppm))	0098	0.64350	0.12100	0.69522	0.15536	0.05881	18	-0.33	038.53	0
038	Molybdenum (mg / kg (ppm))	0964	0.64815	0.04590	0.69522	0.15536	0.05881	18	-0.30	038.43	0
038	Molybdenum (mg / kg (ppm))	0226	0.66500	0.03000	0.69522	0.15536	0.05881	18	-0.19	038.41	0
038	Molybdenum (mg / kg (ppm))	0553	0.71050	0.00500	0.69522	0.15536	0.05881	18	0.10	038.53	0
038	Molybdenum (mg / kg (ppm))	0278	0.73500	0.05000	0.69522	0.15536	0.05881	18	0.26	038.42	0
038	Molybdenum (mg / kg (ppm))	0910	0.76500	0.03000	0.69522	0.15536	0.05881	18	0.45	038.52	0
038	Molybdenum (mg / kg (ppm))	0918	0.77300	0.03360	0.69522	0.15536	0.05881	18	0.50	038.53	0
038	Molybdenum (mg / kg (ppm))	2033	0.78000	0.02000	0.69522	0.15536	0.05881	18	0.55	038.43	0
038	Molybdenum (mg / kg (ppm))	0413	0.78500	0.03000	0.69522	0.15536	0.05881	18	0.58	038.53	0
038	Molybdenum (mg / kg (ppm))	0045	0.86650	0.03300	0.69522	0.15536	0.05881	18	1.10	038.42	0
038	Molybdenum (mg / kg (ppm))	0186	1.0000	0.00000	0.69522	0.15536	0.05881	18	1.96	038.52	0
038	Molybdenum (mg / kg (ppm))	0510	1.0100	0.16000	0.69522	0.15536	0.05881	18	2.03	038.43	0
038	Molybdenum (mg / kg (ppm))	0297	1.1250	0.55000	0.69522	0.15536	0.05881	18	2.77	038.43	1
040	Barium (mg / kg (ppm))	0918	4.8304	0.06580				3		040.53	0
040	Barium (mg / kg (ppm))	0560	5.7050	0.09000				3		040.42	0
040	Barium (mg / kg (ppm))	0508	6.5036	0.16280				3		040.43	0
041	Vanadium (mg / kg (ppm))	0508	2.5942	0.26410				2		041.43	0
041	Vanadium (mg / kg (ppm))	0553	2.7500	0.06000				2		041.53	0
042	Chloride (%)	0638	0.71000	0.00000	0.74075	0.03161	0.01000	4		042.00	0
042	Chloride (%)	0169	0.72500	0.01000	0.74075	0.03161	0.01000	4		042.00	0
042	Chloride (%)	0297	0.74500	0.03000	0.74075	0.03161	0.01000	4		042.00	0
042	Chloride (%)	0948	0.78300	0.00000	0.74075	0.03161	0.01000	4		042.00	0
101	Choline Chloride (mg / kg (ppm))	0675	4,096.3	106.33	5,550.1	988.61	421.58	4		101.00	0
101	Choline Chloride (mg / kg (ppm))	0939	5,763.0	468.00	5,550.1	988.61	421.58	4		101.01	0
101	Choline Chloride (mg / kg (ppm))	0227	6,130.0	60.000	5,550.1	988.61	421.58	4		101.00	0
101	Choline Chloride (mg / kg (ppm))	0051	6,211.0	1,052.0	5,550.1	988.61	421.58	4		101.02	0
102	Niacin (mg / kg (ppm))	0227	250.50	13.000				1		102.01	0
103	Pantothenic Acid (mg / kg (ppm))	0629	1.3800	0.00000				2		103.02	0
103	Pantothenic Acid (mg / kg (ppm))	0227	123.50	7.0000				2		103.01	0
104	Riboflavin (mg / kg (ppm))	0910	63.825	6.2500				3		104.03	0
104	Riboflavin (mg / kg (ppm))	0171	81.500	3.0000				3		104.00	0
104	Riboflavin (mg / kg (ppm))	0227	90.150	0.30000				3		104.00	0
105	Thiamine (mg / kg (ppm))	0629	22.200	0.20000				3		105.00	0
105	Thiamine (mg / kg (ppm))	0227	24.800	2.2000				3		105.01	0
105	Thiamine (mg / kg (ppm))	0910	38.270	4.0200				3		105.00	0
106	Vitamin A (KU / kg)	0227	3.5800	0.02000	5.5147	1.3017	0.72240	21	-1.49	106.02	0
106	Vitamin A (KU / kg)	0905	3.7990	1.2940	5.5147	1.3017	0.72240	21	-1.32	106.02	0
106	Vitamin A (KU / kg)	0861	4.1650	0.61000	5.5147	1.3017	0.72240	21	-1.04	106.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
106	Vitamin A (KU / kg)	0171	4.1850	0.09000	5.5147	1.3017	0.72240	21	-1.02	106.00	0
106	Vitamin A (KU / kg)	2053	4.2000	1.4000	5.5147	1.3017	0.72240	21	-1.01	106.02	0
106	Vitamin A (KU / kg)	2103	4.5795	0.49500	5.5147	1.3017	0.72240	21	-0.72	106.02	0
106	Vitamin A (KU / kg)	0910	4.9000	0.40000	5.5147	1.3017	0.72240	21	-0.47	106.02	0
106	Vitamin A (KU / kg)	0512	4.9990	0.13200	5.5147	1.3017	0.72240	21	-0.40	106.02	0
106	Vitamin A (KU / kg)	0051	5.3915	0.41700	5.5147	1.3017	0.72240	21	-0.09	106.02	0
106	Vitamin A (KU / kg)	0638	5.5350	0.75000	5.5147	1.3017	0.72240	21	0.02	106.02	0
106	Vitamin A (KU / kg)	0098	5.6100	0.18000	5.5147	1.3017	0.72240	21	0.07	106.01	0
106	Vitamin A (KU / kg)	0918	5.8000	0.46000	5.5147	1.3017	0.72240	21	0.22	106.02	0
106	Vitamin A (KU / kg)	0676	5.9050	0.17000	5.5147	1.3017	0.72240	21	0.30	106.02	0
106	Vitamin A (KU / kg)	0017	5.9419	3.0778	5.5147	1.3017	0.72240	21	0.33	106.02	0
106	Vitamin A (KU / kg)	0610	6.1000	0.80000	5.5147	1.3017	0.72240	21	0.45	106.02	0
106	Vitamin A (KU / kg)	2066	6.3650	0.21000	5.5147	1.3017	0.72240	21	0.65	106.02	0
106	Vitamin A (KU / kg)	0148	6.3900	1.1800	5.5147	1.3017	0.72240	21	0.67	106.02	0
106	Vitamin A (KU / kg)	0169	6.5500	0.30000	5.5147	1.3017	0.72240	21	0.80	106.02	0
106	Vitamin A (KU / kg)	0208	6.8785	0.39700	5.5147	1.3017	0.72240	21	1.05	106.02	0
106	Vitamin A (KU / kg)	0027	7.9214	0.12750	5.5147	1.3017	0.72240	21	1.85	106.02	0
106	Vitamin A (KU / kg)	0038	8.4300	2.6600	5.5147	1.3017	0.72240	21	2.24	106.02	0
106	Vitamin A (KU / kg)	0675	38.740	9.6400	5.5147	1.3017	0.72240	21	25.52	106.02	2
107	Vitamin B12 (µg / kg (ppb))	0227	419.00	14.000				1		107.00	0
108	Vitamin D3 (KU / kg)	0861	1.7600	0.00000	3.8617	3.5134	0.26571	7	-0.60	108.02	0
108	Vitamin D3 (KU / kg)	0208	1.8000	0.02000	3.8617	3.5134	0.26571	7	-0.59	108.02	0
108	Vitamin D3 (KU / kg)	0918	1.9550	0.07000	3.8617	3.5134	0.26571	7	-0.54	108.02	0
108	Vitamin D3 (KU / kg)	2004	2.2750	0.05000	3.8617	3.5134	0.26571	7	-0.45	108.99	0
108	Vitamin D3 (KU / kg)	0227	2.5900	0.02000	3.8617	3.5134	0.26571	7	-0.36	108.02	0
108	Vitamin D3 (KU / kg)	0675	7.5200	0.70000	3.8617	3.5134	0.26571	7	1.04	108.02	0
108	Vitamin D3 (KU / kg)	0169	21.500	1.0000	3.8617	3.5134	0.26571	7	5.02	108.01	0
108	Vitamin D3 (KU / kg)	0676	0.00000	0.00000	3.8617	3.5134	0.26571	7		108.02	4
109	Vitamin E (IU/kg)	0619	0.23000	0.00000	902.39	223.75	33.771	19	-4.03	109.02	0
109	Vitamin E (IU/kg)	0590	585.50	33.000	902.39	223.75	33.771	19	-1.42	109.02	0
109	Vitamin E (IU/kg)	0955	623.00	26.000	902.39	223.75	33.771	19	-1.25	109.02	0
109	Vitamin E (IU/kg)	0169	623.50	19.000	902.39	223.75	33.771	19	-1.25	109.02	0
109	Vitamin E (IU/kg)	2053	701.50	11.000	902.39	223.75	33.771	19	-0.90	109.02	0
109	Vitamin E (IU/kg)	0675	753.32	86.110	902.39	223.75	33.771	19	-0.67	109.02	0
109	Vitamin E (IU/kg)	0905	926.00	52.000	902.39	223.75	33.771	19	0.11	109.02	0
109	Vitamin E (IU/kg)	0610	945.00	50.000	902.39	223.75	33.771	19	0.19	109.02	0
109	Vitamin E (IU/kg)	0910	950.75	8.7000	902.39	223.75	33.771	19	0.22	109.02	0
109	Vitamin E (IU/kg)	0227	955.50	65.000	902.39	223.75	33.771	19	0.24	109.02	0
109	Vitamin E (IU/kg)	0861	980.00	20.000	902.39	223.75	33.771	19	0.35	109.02	0
109	Vitamin E (IU/kg)	0171	1,001.3	5.5000	902.39	223.75	33.771	19	0.44	109.99	0
109	Vitamin E (IU/kg)	0918	1,010.5	51.000	902.39	223.75	33.771	19	0.48	109.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
109	Vitamin E (IU/kg)	0098	1,015.0	30.000	902.39	223.75	33.771	19	0.50	109.02	0
109	Vitamin E (IU/kg)	0027	1,025.5	6.2000	902.39	223.75	33.771	19	0.55	109.02	0
109	Vitamin E (IU/kg)	2103	1,032.5	73.000	902.39	223.75	33.771	19	0.58	109.02	0
109	Vitamin E (IU/kg)	0676	1,104.0	34.000	902.39	223.75	33.771	19	0.90	109.02	0
109	Vitamin E (IU/kg)	0148	1,137.9	31.140	902.39	223.75	33.771	19	1.05	109.02	0
109	Vitamin E (IU/kg)	0638	1,208.0	40.000	902.39	223.75	33.771	19	1.37	109.02	0
109	Vitamin E (IU/kg)	0208	705.63	301.57	902.39	223.75	33.771	19	-0.88	109.02	1
111	Vitamin C, phosphorylated (mg / kg (ppn	0861	27.350	0.50000				3		111.01	0
111	Vitamin C, phosphorylated (mg / kg (ppn	0227	38.250	5.1000				3		111.01	0
111	Vitamin C, phosphorylated (mg / kg (ppn	0148	43.100	0.80000				3		111.01	0
112	Pyridoxine (µg / g)	0227	36.400	1.8000				1		112.01	0
113	Folic Acid (mg / kg (ppm))	0227	13.350	3.1000				1		113.01	0
114	Biotin (mg / kg (ppm))	0227	1.0850	0.03000				1		114.01	0
120	Alanine (%)	0226	2.7655	0.17700	3.1127	0.18193	0.04938	27	-1.91	120.00	0
120	Alanine (%)	0297	2.7980	0.05400	3.1127	0.18193	0.04938	27	-1.73	120.01	0
120	Alanine (%)	0626	2.8380	0.10200	3.1127	0.18193	0.04938	27	-1.51	120.05	0
120	Alanine (%)	0148	2.9050	0.11000	3.1127	0.18193	0.04938	27	-1.14	120.05	0
120	Alanine (%)	2144	2.9450	0.05000	3.1127	0.18193	0.04938	27	-0.92	120.99	0
120	Alanine (%)	0610	2.9900	0.04000	3.1127	0.18193	0.04938	27	-0.67	120.05	0
120	Alanine (%)	0910	3.0000	0.06000	3.1127	0.18193	0.04938	27	-0.62	120.00	0
120	Alanine (%)	0098	3.0470	0.14200	3.1127	0.18193	0.04938	27	-0.36	120.02	0
120	Alanine (%)	0644	3.0525	0.00300	3.1127	0.18193	0.04938	27	-0.33	120.00	0
120	Alanine (%)	0878	3.0780	0.00600	3.1127	0.18193	0.04938	27	-0.19	120.00	0
120	Alanine (%)	0859	3.0936	0.00120	3.1127	0.18193	0.04938	27	-0.11	120.00	0
120	Alanine (%)	0504	3.1050	0.07000	3.1127	0.18193	0.04938	27	-0.04	120.00	0
120	Alanine (%)	0652	3.1100	0.08000	3.1127	0.18193	0.04938	27	-0.01	120.00	0
120	Alanine (%)	0968	3.1105	0.03500	3.1127	0.18193	0.04938	27	-0.01	120.00	0
120	Alanine (%)	0354	3.1165	0.02900	3.1127	0.18193	0.04938	27	0.02	120.00	0
120	Alanine (%)	0872	3.1270	0.08600	3.1127	0.18193	0.04938	27	0.08	120.00	0
120	Alanine (%)	0676	3.1315	0.02700	3.1127	0.18193	0.04938	27	0.10	120.05	0
120	Alanine (%)	0870	3.1389	0.01910	3.1127	0.18193	0.04938	27	0.14	120.00	0
120	Alanine (%)	0571	3.1425	0.01300	3.1127	0.18193	0.04938	27	0.16	120.00	0
120	Alanine (%)	2059	3.1820	0.00400	3.1127	0.18193	0.04938	27	0.38	120.00	0
120	Alanine (%)	0675	3.2150	0.01000	3.1127	0.18193	0.04938	27	0.56	120.00	0
120	Alanine (%)	0684	3.2570	0.04800	3.1127	0.18193	0.04938	27	0.79	120.00	0
120	Alanine (%)	0619	3.3000	0.04000	3.1127	0.18193	0.04938	27	1.03	120.00	0
120	Alanine (%)	0227	3.3200	0.00000	3.1127	0.18193	0.04938	27	1.14	120.00	0
120	Alanine (%)	0171	3.4070	0.09200	3.1127	0.18193	0.04938	27	1.62	120.00	0
120	Alanine (%)	0918	3.5212	0.02560	3.1127	0.18193	0.04938	27	2.25	120.00	0
120	Alanine (%)	0889	3.9902	0.00940	3.1127	0.18193	0.04938	27	4.82	120.99	0
121	Arginine (%)	0675	2.4800	0.04000	2.8231	0.10507	0.04464	25	-3.27	121.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
121	Arginine (%)	0226	2.5765	0.00300	2.8231	0.10507	0.04464	25	-2.35	121.00	0
121	Arginine (%)	0878	2.6965	0.03300	2.8231	0.10507	0.04464	25	-1.20	121.00	0
121	Arginine (%)	0870	2.7139	0.00030	2.8231	0.10507	0.04464	25	-1.04	121.00	0
121	Arginine (%)	0626	2.7140	0.06200	2.8231	0.10507	0.04464	25	-1.04	121.05	0
121	Arginine (%)	0859	2.7445	0.00680	2.8231	0.10507	0.04464	25	-0.75	121.00	0
121	Arginine (%)	0910	2.7450	0.07000	2.8231	0.10507	0.04464	25	-0.74	121.00	0
121	Arginine (%)	0148	2.7500	0.24000	2.8231	0.10507	0.04464	25	-0.70	121.05	0
121	Arginine (%)	0098	2.7815	0.00100	2.8231	0.10507	0.04464	25	-0.40	121.02	0
121	Arginine (%)	0652	2.8100	0.06000	2.8231	0.10507	0.04464	25	-0.12	121.00	0
121	Arginine (%)	0504	2.8250	0.03000	2.8231	0.10507	0.04464	25	0.02	121.00	0
121	Arginine (%)	0872	2.8290	0.07600	2.8231	0.10507	0.04464	25	0.06	121.00	0
121	Arginine (%)	0354	2.8310	0.04000	2.8231	0.10507	0.04464	25	0.08	121.00	0
121	Arginine (%)	0684	2.8350	0.04400	2.8231	0.10507	0.04464	25	0.11	121.00	0
121	Arginine (%)	0889	2.8360	0.15080	2.8231	0.10507	0.04464	25	0.12	121.99	0
121	Arginine (%)	0644	2.8615	0.00100	2.8231	0.10507	0.04464	25	0.37	121.00	0
121	Arginine (%)	2059	2.8715	0.00300	2.8231	0.10507	0.04464	25	0.46	121.00	0
121	Arginine (%)	0968	2.8910	0.00000	2.8231	0.10507	0.04464	25	0.65	121.00	0
121	Arginine (%)	0297	2.8990	0.01600	2.8231	0.10507	0.04464	25	0.72	121.01	0
121	Arginine (%)	0171	2.9030	0.06600	2.8231	0.10507	0.04464	25	0.76	121.00	0
121	Arginine (%)	0610	2.9050	0.05000	2.8231	0.10507	0.04464	25	0.78	121.05	0
121	Arginine (%)	0571	2.9070	0.01400	2.8231	0.10507	0.04464	25	0.80	121.00	0
121	Arginine (%)	0227	2.9350	0.05000	2.8231	0.10507	0.04464	25	1.07	121.00	0
121	Arginine (%)	0676	2.9835	0.00900	2.8231	0.10507	0.04464	25	1.53	121.05	0
121	Arginine (%)	0619	3.0450	0.05000	2.8231	0.10507	0.04464	25	2.11	121.00	0
121	Arginine (%)	0918	2.9711	0.27310	2.8231	0.10507	0.04464	25	1.41	121.00	1
122	Aspartic (%)	2144	3.4600	0.26000	3.9617	0.15530	0.07931	27	-3.23	122.99	0
122	Aspartic (%)	0626	3.7095	0.04500	3.9617	0.15530	0.07931	27	-1.62	122.05	0
122	Aspartic (%)	0226	3.7525	0.21900	3.9617	0.15530	0.07931	27	-1.35	122.00	0
122	Aspartic (%)	0910	3.7850	0.07000	3.9617	0.15530	0.07931	27	-1.14	122.00	0
122	Aspartic (%)	0504	3.8350	0.07000	3.9617	0.15530	0.07931	27	-0.82	122.00	0
122	Aspartic (%)	0859	3.8408	0.01730	3.9617	0.15530	0.07931	27	-0.78	122.00	0
122	Aspartic (%)	0297	3.8470	0.11200	3.9617	0.15530	0.07931	27	-0.74	122.01	0
122	Aspartic (%)	0354	3.8890	0.01800	3.9617	0.15530	0.07931	27	-0.47	122.00	0
122	Aspartic (%)	0675	3.9300	0.02000	3.9617	0.15530	0.07931	27	-0.20	122.00	0
122	Aspartic (%)	0148	3.9350	0.31000	3.9617	0.15530	0.07931	27	-0.17	122.05	0
122	Aspartic (%)	0610	3.9350	0.07000	3.9617	0.15530	0.07931	27	-0.17	122.05	0
122	Aspartic (%)	0644	3.9385	0.02500	3.9617	0.15530	0.07931	27	-0.15	122.00	0
122	Aspartic (%)	0870	3.9527	0.01000	3.9617	0.15530	0.07931	27	-0.06	122.00	0
122	Aspartic (%)	0872	3.9610	0.11800	3.9617	0.15530	0.07931	27	0.00	122.00	0
122	Aspartic (%)	0684	3.9680	0.07000	3.9617	0.15530	0.07931	27	0.04	122.00	0
122	Aspartic (%)	0571	3.9745	0.03900	3.9617	0.15530	0.07931	27	0.08	122.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
122	Aspartic (%)	0652	3.9750	0.15000	3.9617	0.15530	0.07931	27	0.09	122.00	0
122	Aspartic (%)	0098	3.9770	0.01800	3.9617	0.15530	0.07931	27	0.10	122.02	0
122	Aspartic (%)	0968	3.9860	0.00400	3.9617	0.15530	0.07931	27	0.16	122.00	0
122	Aspartic (%)	2059	4.0385	0.00300	3.9617	0.15530	0.07931	27	0.49	122.00	0
122	Aspartic (%)	0878	4.0530	0.06000	3.9617	0.15530	0.07931	27	0.59	122.00	0
122	Aspartic (%)	0227	4.0750	0.03000	3.9617	0.15530	0.07931	27	0.73	122.00	0
122	Aspartic (%)	0676	4.1320	0.11800	3.9617	0.15530	0.07931	27	1.10	122.05	0
122	Aspartic (%)	0619	4.1650	0.07000	3.9617	0.15530	0.07931	27	1.31	122.00	0
122	Aspartic (%)	0889	4.1740	0.03770	3.9617	0.15530	0.07931	27	1.37	122.99	0
122	Aspartic (%)	0918	4.2198	0.16050	3.9617	0.15530	0.07931	27	1.66	122.00	0
122	Aspartic (%)	0171	4.2735	0.01700	3.9617	0.15530	0.07931	27	2.01	122.00	0
124	Cysteine/Cystine (%)	0889	0.39100	0.06600	0.64660	0.04284	0.01990	26	-5.97	124.99	0
124	Cysteine/Cystine (%)	0870	0.58430	0.03280	0.64660	0.04284	0.01990	26	-1.45	124.00	0
124	Cysteine/Cystine (%)	0227	0.60000	0.00000	0.64660	0.04284	0.01990	26	-1.09	124.00	0
124	Cysteine/Cystine (%)	0644	0.61200	0.01400	0.64660	0.04284	0.01990	26	-0.81	124.00	0
124	Cysteine/Cystine (%)	0910	0.61500	0.01000	0.64660	0.04284	0.01990	26	-0.74	124.00	0
124	Cysteine/Cystine (%)	0148	0.61500	0.03000	0.64660	0.04284	0.01990	26	-0.74	124.05	0
124	Cysteine/Cystine (%)	0859	0.62210	0.00160	0.64660	0.04284	0.01990	26	-0.57	124.00	0
124	Cysteine/Cystine (%)	0226	0.62650	0.06500	0.64660	0.04284	0.01990	26	-0.47	124.00	0
124	Cysteine/Cystine (%)	0676	0.63050	0.03300	0.64660	0.04284	0.01990	26	-0.38	124.05	0
124	Cysteine/Cystine (%)	0619	0.63150	0.01900	0.64660	0.04284	0.01990	26	-0.35	124.00	0
124	Cysteine/Cystine (%)	0918	0.63340	0.00300	0.64660	0.04284	0.01990	26	-0.31	124.00	0
124	Cysteine/Cystine (%)	0652	0.64000	0.02000	0.64660	0.04284	0.01990	26	-0.15	124.00	0
124	Cysteine/Cystine (%)	0878	0.64100	0.01000	0.64660	0.04284	0.01990	26	-0.13	124.00	0
124	Cysteine/Cystine (%)	0098	0.64150	0.00700	0.64660	0.04284	0.01990	26	-0.12	124.02	0
124	Cysteine/Cystine (%)	0571	0.64200	0.00000	0.64660	0.04284	0.01990	26	-0.11	124.00	0
124	Cysteine/Cystine (%)	0684	0.64300	0.00600	0.64660	0.04284	0.01990	26	-0.08	124.00	0
124	Cysteine/Cystine (%)	0872	0.65300	0.02600	0.64660	0.04284	0.01990	26	0.15	124.00	0
124	Cysteine/Cystine (%)	2059	0.65600	0.00600	0.64660	0.04284	0.01990	26	0.22	124.00	0
124	Cysteine/Cystine (%)	0968	0.65850	0.00300	0.64660	0.04284	0.01990	26	0.28	124.00	0
124	Cysteine/Cystine (%)	0297	0.67100	0.04400	0.64660	0.04284	0.01990	26	0.57	124.01	0
124	Cysteine/Cystine (%)	0171	0.67450	0.03100	0.64660	0.04284	0.01990	26	0.65	124.00	0
124	Cysteine/Cystine (%)	0610	0.69500	0.01000	0.64660	0.04284	0.01990	26	1.13	124.05	0
124	Cysteine/Cystine (%)	0354	0.71200	0.01000	0.64660	0.04284	0.01990	26	1.53	124.00	0
124	Cysteine/Cystine (%)	0504	0.74500	0.01000	0.64660	0.04284	0.01990	26	2.30	124.00	0
124	Cysteine/Cystine (%)	0675	0.82000	0.04000	0.64660	0.04284	0.01990	26	4.05	124.00	0
124	Cysteine/Cystine (%)	2144	0.88000	0.02000	0.64660	0.04284	0.01990	26	5.45	124.99	0
125	Glutamic (%)	2144	6.1900	0.06000	6.9366	0.28860	0.10564	27	-2.59	125.99	0
125	Glutamic (%)	0171	6.2135	0.07700	6.9366	0.28860	0.10564	27	-2.51	125.00	0
125	Glutamic (%)	0626	6.5510	0.01000	6.9366	0.28860	0.10564	27	-1.34	125.05	0
125	Glutamic (%)	0878	6.6620	0.10000	6.9366	0.28860	0.10564	27	-0.95	125.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
125	Glutamic (%)	0504	6.7200	0.18000	6.9366	0.28860	0.10564	27	-0.75	125.00	0
125	Glutamic (%)	0859	6.7379	0.00690	6.9366	0.28860	0.10564	27	-0.69	125.00	0
125	Glutamic (%)	0354	6.7665	0.04100	6.9366	0.28860	0.10564	27	-0.59	125.00	0
125	Glutamic (%)	0910	6.7850	0.13000	6.9366	0.28860	0.10564	27	-0.53	125.00	0
125	Glutamic (%)	0098	6.7940	0.03000	6.9366	0.28860	0.10564	27	-0.49	125.02	0
125	Glutamic (%)	0644	6.8635	0.01500	6.9366	0.28860	0.10564	27	-0.25	125.00	0
125	Glutamic (%)	0968	6.8775	0.00100	6.9366	0.28860	0.10564	27	-0.20	125.00	0
125	Glutamic (%)	0872	6.8815	0.20100	6.9366	0.28860	0.10564	27	-0.19	125.00	0
125	Glutamic (%)	0610	6.8950	0.13000	6.9366	0.28860	0.10564	27	-0.14	125.05	0
125	Glutamic (%)	0571	6.9045	0.02300	6.9366	0.28860	0.10564	27	-0.11	125.00	0
125	Glutamic (%)	0652	6.9050	0.23000	6.9366	0.28860	0.10564	27	-0.11	125.00	0
125	Glutamic (%)	0684	6.9260	0.12200	6.9366	0.28860	0.10564	27	-0.04	125.00	0
125	Glutamic (%)	2059	6.9630	0.01400	6.9366	0.28860	0.10564	27	0.09	125.00	0
125	Glutamic (%)	0226	6.9825	0.38700	6.9366	0.28860	0.10564	27	0.16	125.00	0
125	Glutamic (%)	0675	7.0150	0.05000	6.9366	0.28860	0.10564	27	0.27	125.00	0
125	Glutamic (%)	0870	7.0287	0.06740	6.9366	0.28860	0.10564	27	0.32	125.00	0
125	Glutamic (%)	0227	7.2100	0.00000	6.9366	0.28860	0.10564	27	0.95	125.00	0
125	Glutamic (%)	0297	7.2320	0.08400	6.9366	0.28860	0.10564	27	1.02	125.01	0
125	Glutamic (%)	0148	7.2400	0.40000	6.9366	0.28860	0.10564	27	1.05	125.05	0
125	Glutamic (%)	0918	7.2940	0.26820	6.9366	0.28860	0.10564	27	1.24	125.00	0
125	Glutamic (%)	0676	7.3070	0.04600	6.9366	0.28860	0.10564	27	1.28	125.05	0
125	Glutamic (%)	0619	7.5100	0.16000	6.9366	0.28860	0.10564	27	1.99	125.00	0
125	Glutamic (%)	0889	8.0558	0.01880	6.9366	0.28860	0.10564	27	3.88	125.99	0
126	Glycine (%)	0148	3.4350	0.15000	3.6771	0.10733	0.05135	26	-2.26	126.05	0
126	Glycine (%)	2144	3.4400	0.02000	3.6771	0.10733	0.05135	26	-2.21	126.99	0
126	Glycine (%)	0226	3.4945	0.15100	3.6771	0.10733	0.05135	26	-1.70	126.00	0
126	Glycine (%)	0910	3.5300	0.02000	3.6771	0.10733	0.05135	26	-1.37	126.00	0
126	Glycine (%)	0652	3.5650	0.09000	3.6771	0.10733	0.05135	26	-1.04	126.00	0
126	Glycine (%)	0626	3.6175	0.03100	3.6771	0.10733	0.05135	26	-0.56	126.05	0
126	Glycine (%)	0676	3.6225	0.07300	3.6771	0.10733	0.05135	26	-0.51	126.05	0
126	Glycine (%)	0859	3.6295	0.01080	3.6771	0.10733	0.05135	26	-0.44	126.00	0
126	Glycine (%)	0504	3.6400	0.04000	3.6771	0.10733	0.05135	26	-0.35	126.00	0
126	Glycine (%)	0870	3.6507	0.00590	3.6771	0.10733	0.05135	26	-0.25	126.00	0
126	Glycine (%)	0354	3.6545	0.04500	3.6771	0.10733	0.05135	26	-0.21	126.00	0
126	Glycine (%)	0098	3.6810	0.00000	3.6771	0.10733	0.05135	26	0.04	126.02	0
126	Glycine (%)	0297	3.6840	0.15000	3.6771	0.10733	0.05135	26	0.06	126.01	0
126	Glycine (%)	0872	3.6930	0.08400	3.6771	0.10733	0.05135	26	0.15	126.00	0
126	Glycine (%)	0968	3.6980	0.01800	3.6771	0.10733	0.05135	26	0.19	126.00	0
126	Glycine (%)	0610	3.7100	0.08000	3.6771	0.10733	0.05135	26	0.31	126.05	0
126	Glycine (%)	0171	3.7305	0.06500	3.6771	0.10733	0.05135	26	0.50	126.00	0
126	Glycine (%)	0571	3.7335	0.00100	3.6771	0.10733	0.05135	26	0.53	126.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
126	Glycine (%)	0684	3.7380	0.12200	3.6771	0.10733	0.05135	26	0.57	126.00	0
126	Glycine (%)	0644	3.7430	0.00400	3.6771	0.10733	0.05135	26	0.61	126.00	0
126	Glycine (%)	0675	3.7450	0.01000	3.6771	0.10733	0.05135	26	0.63	126.00	0
126	Glycine (%)	2059	3.7510	0.00200	3.6771	0.10733	0.05135	26	0.69	126.00	0
126	Glycine (%)	0878	3.7565	0.08500	3.6771	0.10733	0.05135	26	0.74	126.00	0
126	Glycine (%)	0619	3.8150	0.03000	3.6771	0.10733	0.05135	26	1.28	126.00	0
126	Glycine (%)	0227	3.8300	0.02000	3.6771	0.10733	0.05135	26	1.42	126.00	0
126	Glycine (%)	0918	4.1547	0.02740	3.6771	0.10733	0.05135	26	4.45	126.00	0
126	Glycine (%)	0889	2.3932	0.52760	3.6771	0.10733	0.05135	26	-11.96	126.99	1
127	Histidine (%)	0889	0.83860	0.00000	1.1343	0.05805	0.02325	26	-5.09	127.99	0
127	Histidine (%)	0226	0.98800	0.05000	1.1343	0.05805	0.02325	26	-2.52	127.00	0
127	Histidine (%)	0297	1.0230	0.04000	1.1343	0.05805	0.02325	26	-1.92	127.01	0
127	Histidine (%)	0870	1.0829	0.00080	1.1343	0.05805	0.02325	26	-0.88	127.00	0
127	Histidine (%)	0859	1.0851	0.00800	1.1343	0.05805	0.02325	26	-0.85	127.00	0
127	Histidine (%)	0910	1.0900	0.04000	1.1343	0.05805	0.02325	26	-0.76	127.00	0
127	Histidine (%)	0098	1.0935	0.00700	1.1343	0.05805	0.02325	26	-0.70	127.02	0
127	Histidine (%)	0626	1.0985	0.01500	1.1343	0.05805	0.02325	26	-0.62	127.05	0
127	Histidine (%)	0684	1.1200	0.02400	1.1343	0.05805	0.02325	26	-0.25	127.00	0
127	Histidine (%)	2059	1.1265	0.00300	1.1343	0.05805	0.02325	26	-0.13	127.00	0
127	Histidine (%)	0571	1.1270	0.01600	1.1343	0.05805	0.02325	26	-0.13	127.00	0
127	Histidine (%)	0610	1.1300	0.02000	1.1343	0.05805	0.02325	26	-0.07	127.05	0
127	Histidine (%)	0676	1.1335	0.00500	1.1343	0.05805	0.02325	26	-0.01	127.05	0
127	Histidine (%)	0968	1.1350	0.00800	1.1343	0.05805	0.02325	26	0.01	127.00	0
127	Histidine (%)	0148	1.1400	0.08000	1.1343	0.05805	0.02325	26	0.10	127.05	0
127	Histidine (%)	0644	1.1425	0.00100	1.1343	0.05805	0.02325	26	0.14	127.00	0
127	Histidine (%)	0354	1.1470	0.01200	1.1343	0.05805	0.02325	26	0.22	127.00	0
127	Histidine (%)	0504	1.1550	0.07000	1.1343	0.05805	0.02325	26	0.36	127.00	0
127	Histidine (%)	0675	1.1600	0.00000	1.1343	0.05805	0.02325	26	0.44	127.00	0
127	Histidine (%)	0227	1.1650	0.01000	1.1343	0.05805	0.02325	26	0.53	127.00	0
127	Histidine (%)	0652	1.1800	0.06000	1.1343	0.05805	0.02325	26	0.79	127.00	0
127	Histidine (%)	0872	1.1825	0.03300	1.1343	0.05805	0.02325	26	0.83	127.00	0
127	Histidine (%)	0878	1.1955	0.03300	1.1343	0.05805	0.02325	26	1.05	127.00	0
127	Histidine (%)	0171	1.2170	0.05800	1.1343	0.05805	0.02325	26	1.43	127.00	0
127	Histidine (%)	0918	1.2407	0.00080	1.1343	0.05805	0.02325	26	1.83	127.00	0
127	Histidine (%)	0619	1.3050	0.01000	1.1343	0.05805	0.02325	26	2.94	127.00	0
128	Isoleucine (%)	0226	1.6400	0.16000	1.8484	0.07916	0.04203	27	-2.63	128.00	0
128	Isoleucine (%)	0859	1.7388	0.00730	1.8484	0.07916	0.04203	27	-1.39	128.00	0
128	Isoleucine (%)	0889	1.7572	0.17900	1.8484	0.07916	0.04203	27	-1.15	128.99	0
128	Isoleucine (%)	0652	1.7650	0.03000	1.8484	0.07916	0.04203	27	-1.05	128.00	0
128	Isoleucine (%)	0910	1.7700	0.02000	1.8484	0.07916	0.04203	27	-0.99	128.00	0
128	Isoleucine (%)	0676	1.7820	0.01200	1.8484	0.07916	0.04203	27	-0.84	128.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
128	Isoleucine (%)	2144	1.8050	0.15000	1.8484	0.07916	0.04203	27	-0.55	128.99	0
128	Isoleucine (%)	0684	1.8095	0.02300	1.8484	0.07916	0.04203	27	-0.49	128.00	0
128	Isoleucine (%)	0878	1.8125	0.00300	1.8484	0.07916	0.04203	27	-0.45	128.00	0
128	Isoleucine (%)	0098	1.8130	0.02800	1.8484	0.07916	0.04203	27	-0.45	128.02	0
128	Isoleucine (%)	2059	1.8165	0.00100	1.8484	0.07916	0.04203	27	-0.40	128.00	0
128	Isoleucine (%)	0968	1.8270	0.00600	1.8484	0.07916	0.04203	27	-0.27	128.00	0
128	Isoleucine (%)	0872	1.8305	0.04900	1.8484	0.07916	0.04203	27	-0.23	128.00	0
128	Isoleucine (%)	0918	1.8481	0.15370	1.8484	0.07916	0.04203	27	-0.01	128.00	0
128	Isoleucine (%)	0354	1.8535	0.00700	1.8484	0.07916	0.04203	27	0.06	128.00	0
128	Isoleucine (%)	0571	1.8620	0.01600	1.8484	0.07916	0.04203	27	0.17	128.00	0
128	Isoleucine (%)	0171	1.8680	0.05600	1.8484	0.07916	0.04203	27	0.25	128.00	0
128	Isoleucine (%)	0610	1.8700	0.04000	1.8484	0.07916	0.04203	27	0.27	128.05	0
128	Isoleucine (%)	0619	1.8800	0.04000	1.8484	0.07916	0.04203	27	0.40	128.00	0
128	Isoleucine (%)	0675	1.8850	0.03000	1.8484	0.07916	0.04203	27	0.46	128.00	0
128	Isoleucine (%)	0644	1.9075	0.00700	1.8484	0.07916	0.04203	27	0.75	128.00	0
128	Isoleucine (%)	0626	1.9080	0.00200	1.8484	0.07916	0.04203	27	0.75	128.05	0
128	Isoleucine (%)	0227	1.9300	0.00000	1.8484	0.07916	0.04203	27	1.03	128.00	0
128	Isoleucine (%)	0504	1.9500	0.02000	1.8484	0.07916	0.04203	27	1.28	128.00	0
128	Isoleucine (%)	0148	1.9550	0.01000	1.8484	0.07916	0.04203	27	1.35	128.05	0
128	Isoleucine (%)	0297	1.9930	0.00200	1.8484	0.07916	0.04203	27	1.83	128.01	0
128	Isoleucine (%)	0870	2.0170	0.08270	1.8484	0.07916	0.04203	27	2.13	128.00	0
129	Leucine (%)	0910	3.7950	0.03000	4.0047	0.09711	0.04535	27	-2.16	129.00	0
129	Leucine (%)	0870	3.8228	0.06050	4.0047	0.09711	0.04535	27	-1.87	129.00	0
129	Leucine (%)	0872	3.9090	0.10800	4.0047	0.09711	0.04535	27	-0.99	129.00	0
129	Leucine (%)	0652	3.9150	0.07000	4.0047	0.09711	0.04535	27	-0.92	129.00	0
129	Leucine (%)	0859	3.9216	0.01430	4.0047	0.09711	0.04535	27	-0.86	129.00	0
129	Leucine (%)	0098	3.9345	0.00300	4.0047	0.09711	0.04535	27	-0.72	129.02	0
129	Leucine (%)	2059	3.9445	0.01900	4.0047	0.09711	0.04535	27	-0.62	129.00	0
129	Leucine (%)	0626	3.9500	0.05000	4.0047	0.09711	0.04535	27	-0.56	129.05	0
129	Leucine (%)	2144	3.9550	0.03000	4.0047	0.09711	0.04535	27	-0.51	129.99	0
129	Leucine (%)	0226	3.9660	0.17800	4.0047	0.09711	0.04535	27	-0.40	129.00	0
129	Leucine (%)	0354	3.9855	0.03500	4.0047	0.09711	0.04535	27	-0.20	129.00	0
129	Leucine (%)	0684	3.9890	0.05200	4.0047	0.09711	0.04535	27	-0.16	129.00	0
129	Leucine (%)	0504	3.9900	0.04000	4.0047	0.09711	0.04535	27	-0.15	129.00	0
129	Leucine (%)	0148	3.9900	0.02000	4.0047	0.09711	0.04535	27	-0.15	129.05	0
129	Leucine (%)	0644	3.9920	0.00800	4.0047	0.09711	0.04535	27	-0.13	129.00	0
129	Leucine (%)	0571	3.9960	0.03200	4.0047	0.09711	0.04535	27	-0.09	129.00	0
129	Leucine (%)	0297	4.0075	0.02900	4.0047	0.09711	0.04535	27	0.03	129.01	0
129	Leucine (%)	0968	4.0375	0.07500	4.0047	0.09711	0.04535	27	0.34	129.00	0
129	Leucine (%)	0227	4.0450	0.01000	4.0047	0.09711	0.04535	27	0.42	129.00	0
129	Leucine (%)	0676	4.0470	0.01200	4.0047	0.09711	0.04535	27	0.44	129.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
129	Leucine (%)	0878	4.0560	0.01200	4.0047	0.09711	0.04535	27	0.53	129.00	0
129	Leucine (%)	0610	4.0600	0.00000	4.0047	0.09711	0.04535	27	0.57	129.05	0
129	Leucine (%)	0171	4.1155	0.03100	4.0047	0.09711	0.04535	27	1.14	129.00	0
129	Leucine (%)	0889	4.1504	0.04710	4.0047	0.09711	0.04535	27	1.50	129.99	0
129	Leucine (%)	0675	4.2200	0.04000	4.0047	0.09711	0.04535	27	2.22	129.00	0
129	Leucine (%)	0619	4.2250	0.07000	4.0047	0.09711	0.04535	27	2.27	129.00	0
129	Leucine (%)	0918	4.3414	0.14850	4.0047	0.09711	0.04535	27	3.47	129.00	0
130	L-Lysine (%)	0889	2.5110	0.02830	4.1795	0.23999	0.08037	30	-6.95	130.99	0
130	L-Lysine (%)	0171	3.5935	0.02500	4.1795	0.23999	0.08037	30	-2.44	130.00	0
130	L-Lysine (%)	0226	3.7575	0.19900	4.1795	0.23999	0.08037	30	-1.76	130.00	0
130	L-Lysine (%)	0848	3.8400	0.01000	4.1795	0.23999	0.08037	30	-1.41	130.00	0
130	L-Lysine (%)	0626	3.9135	0.01700	4.1795	0.23999	0.08037	30	-1.11	130.05	0
130	L-Lysine (%)	2053	3.9450	0.19000	4.1795	0.23999	0.08037	30	-0.98	130.99	0
130	L-Lysine (%)	0910	3.9650	0.11000	4.1795	0.23999	0.08037	30	-0.89	130.00	0
130	L-Lysine (%)	0354	4.0235	0.01900	4.1795	0.23999	0.08037	30	-0.65	130.00	0
130	L-Lysine (%)	0652	4.0350	0.05000	4.1795	0.23999	0.08037	30	-0.60	130.00	0
130	L-Lysine (%)	2144	4.1350	0.07000	4.1795	0.23999	0.08037	30	-0.19	130.99	0
130	L-Lysine (%)	0872	4.1400	0.09800	4.1795	0.23999	0.08037	30	-0.16	130.00	0
130	L-Lysine (%)	0610	4.1700	0.14000	4.1795	0.23999	0.08037	30	-0.04	130.05	0
130	L-Lysine (%)	0676	4.1850	0.04400	4.1795	0.23999	0.08037	30	0.02	130.05	0
130	L-Lysine (%)	0684	4.2010	0.11200	4.1795	0.23999	0.08037	30	0.09	130.00	0
130	L-Lysine (%)	0571	4.2140	0.04000	4.1795	0.23999	0.08037	30	0.14	130.00	0
130	L-Lysine (%)	0644	4.2195	0.00900	4.1795	0.23999	0.08037	30	0.17	130.00	0
130	L-Lysine (%)	0878	4.2210	0.02200	4.1795	0.23999	0.08037	30	0.17	130.00	0
130	L-Lysine (%)	0968	4.2405	0.01900	4.1795	0.23999	0.08037	30	0.25	130.00	0
130	L-Lysine (%)	0512	4.2740	0.19000	4.1795	0.23999	0.08037	30	0.39	130.00	0
130	L-Lysine (%)	2059	4.2755	0.00500	4.1795	0.23999	0.08037	30	0.40	130.00	0
130	L-Lysine (%)	0859	4.2787	0.27480	4.1795	0.23999	0.08037	30	0.41	130.00	0
130	L-Lysine (%)	0504	4.2800	0.04000	4.1795	0.23999	0.08037	30	0.42	130.00	0
130	L-Lysine (%)	0675	4.2850	0.05000	4.1795	0.23999	0.08037	30	0.44	130.00	0
130	L-Lysine (%)	0297	4.3060	0.01000	4.1795	0.23999	0.08037	30	0.53	130.01	0
130	L-Lysine (%)	0870	4.3626	0.00470	4.1795	0.23999	0.08037	30	0.76	130.00	0
130	L-Lysine (%)	0098	4.4475	0.02900	4.1795	0.23999	0.08037	30	1.12	130.02	0
130	L-Lysine (%)	0227	4.4600	0.08000	4.1795	0.23999	0.08037	30	1.17	130.00	0
130	L-Lysine (%)	0619	4.4850	0.07000	4.1795	0.23999	0.08037	30	1.27	130.00	0
130	L-Lysine (%)	0148	4.4850	0.31000	4.1795	0.23999	0.08037	30	1.27	130.05	0
130	L-Lysine (%)	0918	4.5561	0.14520	4.1795	0.23999	0.08037	30	1.57	130.00	0
130	L-Lysine (%)	0014	4.0750	0.55000	4.1795	0.23999	0.08037	30	-0.44	130.05	1
131	Methionine (%)	0889	0.96100	0.00000	1.4323	0.09886	0.03327	29	-4.77	131.99	0
131	Methionine (%)	0619	1.2500	0.04000	1.4323	0.09886	0.03327	29	-1.84	131.00	0
131	Methionine (%)	0014	1.2850	0.05000	1.4323	0.09886	0.03327	29	-1.49	131.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
131	Methionine (%)	0878	1.3200	0.02000	1.4323	0.09886	0.03327	29	-1.14	131.00	0
131	Methionine (%)	0684	1.3550	0.05800	1.4323	0.09886	0.03327	29	-0.78	131.00	0
131	Methionine (%)	0098	1.3560	0.02000	1.4323	0.09886	0.03327	29	-0.77	131.02	0
131	Methionine (%)	0859	1.3617	0.00110	1.4323	0.09886	0.03327	29	-0.71	131.00	0
131	Methionine (%)	0968	1.3795	0.00700	1.4323	0.09886	0.03327	29	-0.53	131.00	0
131	Methionine (%)	0297	1.3845	0.00700	1.4323	0.09886	0.03327	29	-0.48	131.01	0
131	Methionine (%)	0918	1.3905	0.04440	1.4323	0.09886	0.03327	29	-0.42	131.00	0
131	Methionine (%)	0910	1.3950	0.03000	1.4323	0.09886	0.03327	29	-0.38	131.00	0
131	Methionine (%)	0226	1.4000	0.05200	1.4323	0.09886	0.03327	29	-0.33	131.00	0
131	Methionine (%)	0504	1.4000	0.02000	1.4323	0.09886	0.03327	29	-0.33	131.00	0
131	Methionine (%)	0148	1.4250	0.11000	1.4323	0.09886	0.03327	29	-0.07	131.05	0
131	Methionine (%)	0652	1.4300	0.02000	1.4323	0.09886	0.03327	29	-0.02	131.00	0
131	Methionine (%)	0171	1.4455	0.09700	1.4323	0.09886	0.03327	29	0.13	131.00	0
131	Methionine (%)	2059	1.4490	0.01400	1.4323	0.09886	0.03327	29	0.17	131.00	0
131	Methionine (%)	0610	1.4500	0.02000	1.4323	0.09886	0.03327	29	0.18	131.05	0
131	Methionine (%)	0870	1.4576	0.03620	1.4323	0.09886	0.03327	29	0.26	131.00	0
131	Methionine (%)	2053	1.4650	0.09000	1.4323	0.09886	0.03327	29	0.33	131.99	0
131	Methionine (%)	0571	1.4815	0.00100	1.4323	0.09886	0.03327	29	0.50	131.00	0
131	Methionine (%)	0644	1.5015	0.00700	1.4323	0.09886	0.03327	29	0.70	131.00	0
131	Methionine (%)	0872	1.5195	0.03500	1.4323	0.09886	0.03327	29	0.88	131.00	0
131	Methionine (%)	0626	1.5240	0.00400	1.4323	0.09886	0.03327	29	0.93	131.05	0
131	Methionine (%)	0354	1.5305	0.00500	1.4323	0.09886	0.03327	29	0.99	131.00	0
131	Methionine (%)	0227	1.5450	0.03000	1.4323	0.09886	0.03327	29	1.14	131.00	0
131	Methionine (%)	0512	1.5560	0.07600	1.4323	0.09886	0.03327	29	1.25	131.00	0
131	Methionine (%)	2144	1.6450	0.03000	1.4323	0.09886	0.03327	29	2.15	131.99	0
131	Methionine (%)	0675	1.9200	0.04000	1.4323	0.09886	0.03327	29	4.93	131.00	0
132	Phenylalanine (%)	0889	1.8043	0.00940	2.1461	0.09173	0.04181	25	-3.73	132.99	0
132	Phenylalanine (%)	0226	1.8675	0.04300	2.1461	0.09173	0.04181	25	-3.04	132.00	0
132	Phenylalanine (%)	0910	1.9350	0.01000	2.1461	0.09173	0.04181	25	-2.30	132.00	0
132	Phenylalanine (%)	0098	2.0715	0.01500	2.1461	0.09173	0.04181	25	-0.81	132.02	0
132	Phenylalanine (%)	0626	2.0825	0.04100	2.1461	0.09173	0.04181	25	-0.69	132.05	0
132	Phenylalanine (%)	0684	2.0845	0.04500	2.1461	0.09173	0.04181	25	-0.67	132.00	0
132	Phenylalanine (%)	0227	2.0950	0.03000	2.1461	0.09173	0.04181	25	-0.56	132.00	0
132	Phenylalanine (%)	0652	2.1050	0.09000	2.1461	0.09173	0.04181	25	-0.45	132.00	0
132	Phenylalanine (%)	0297	2.1130	0.07800	2.1461	0.09173	0.04181	25	-0.36	132.01	0
132	Phenylalanine (%)	0872	2.1140	0.10800	2.1461	0.09173	0.04181	25	-0.35	132.00	0
132	Phenylalanine (%)	0354	2.1265	0.00500	2.1461	0.09173	0.04181	25	-0.21	132.00	0
132	Phenylalanine (%)	0571	2.1355	0.02900	2.1461	0.09173	0.04181	25	-0.12	132.00	0
132	Phenylalanine (%)	0968	2.1400	0.08000	2.1461	0.09173	0.04181	25	-0.07	132.00	0
132	Phenylalanine (%)	0644	2.1545	0.01700	2.1461	0.09173	0.04181	25	0.09	132.00	0
132	Phenylalanine (%)	0610	2.1550	0.05000	2.1461	0.09173	0.04181	25	0.10	132.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
132	Phenylalanine (%)	0676	2.1585	0.00700	2.1461	0.09173	0.04181	25	0.14	132.05	0
132	Phenylalanine (%)	0504	2.1700	0.04000	2.1461	0.09173	0.04181	25	0.26	132.00	0
132	Phenylalanine (%)	0878	2.1845	0.06700	2.1461	0.09173	0.04181	25	0.42	132.00	0
132	Phenylalanine (%)	2059	2.2200	0.00600	2.1461	0.09173	0.04181	25	0.81	132.00	0
132	Phenylalanine (%)	0171	2.2215	0.05900	2.1461	0.09173	0.04181	25	0.82	132.00	0
132	Phenylalanine (%)	0859	2.2292	0.02010	2.1461	0.09173	0.04181	25	0.91	132.00	0
132	Phenylalanine (%)	0619	2.2350	0.03000	2.1461	0.09173	0.04181	25	0.97	132.00	0
132	Phenylalanine (%)	0870	2.2632	0.08260	2.1461	0.09173	0.04181	25	1.28	132.00	0
132	Phenylalanine (%)	0918	2.2839	0.06310	2.1461	0.09173	0.04181	25	1.50	132.00	0
132	Phenylalanine (%)	0675	2.3200	0.02000	2.1461	0.09173	0.04181	25	1.90	132.00	0
132	Phenylalanine (%)	0148	2.0600	0.20000	2.1461	0.09173	0.04181	25	-0.94	132.05	1
133	Proline (%)	0675	1.9000	0.04000	3.1183	0.13097	0.05777	24	-9.30	133.00	0
133	Proline (%)	0859	2.7089	0.01050	3.1183	0.13097	0.05777	24	-3.13	133.00	0
133	Proline (%)	2144	2.8200	0.14000	3.1183	0.13097	0.05777	24	-2.28	133.99	0
133	Proline (%)	0889	2.9821	0.27330	3.1183	0.13097	0.05777	24	-1.04	133.99	0
133	Proline (%)	0171	3.0330	0.18000	3.1183	0.13097	0.05777	24	-0.65	133.00	0
133	Proline (%)	0910	3.0500	0.08000	3.1183	0.13097	0.05777	24	-0.52	133.00	0
133	Proline (%)	0878	3.0525	0.01500	3.1183	0.13097	0.05777	24	-0.50	133.00	0
133	Proline (%)	0872	3.0620	0.02000	3.1183	0.13097	0.05777	24	-0.43	133.00	0
133	Proline (%)	0148	3.0700	0.02000	3.1183	0.13097	0.05777	24	-0.37	133.05	0
133	Proline (%)	0870	3.0809	0.04460	3.1183	0.13097	0.05777	24	-0.29	133.00	0
133	Proline (%)	0226	3.0815	0.04500	3.1183	0.13097	0.05777	24	-0.28	133.00	0
133	Proline (%)	2059	3.1090	0.00000	3.1183	0.13097	0.05777	24	-0.07	133.00	0
133	Proline (%)	0652	3.1100	0.06000	3.1183	0.13097	0.05777	24	-0.06	133.00	0
133	Proline (%)	0571	3.1590	0.03800	3.1183	0.13097	0.05777	24	0.31	133.00	0
133	Proline (%)	0504	3.1700	0.04000	3.1183	0.13097	0.05777	24	0.39	133.00	0
133	Proline (%)	0610	3.1700	0.06000	3.1183	0.13097	0.05777	24	0.39	133.05	0
133	Proline (%)	0968	3.1850	0.02600	3.1183	0.13097	0.05777	24	0.51	133.00	0
133	Proline (%)	0619	3.1950	0.03000	3.1183	0.13097	0.05777	24	0.59	133.00	0
133	Proline (%)	0354	3.2300	0.00200	3.1183	0.13097	0.05777	24	0.85	133.00	0
133	Proline (%)	0644	3.2305	0.02700	3.1183	0.13097	0.05777	24	0.86	133.00	0
133	Proline (%)	0684	3.2355	0.03900	3.1183	0.13097	0.05777	24	0.89	133.00	0
133	Proline (%)	0626	3.2385	0.04700	3.1183	0.13097	0.05777	24	0.92	133.05	0
133	Proline (%)	0676	3.3395	0.01900	3.1183	0.13097	0.05777	24	1.69	133.05	0
133	Proline (%)	0227	3.4150	0.13000	3.1183	0.13097	0.05777	24	2.27	133.00	0
133	Proline (%)	0918	4.4227	0.33920	3.1183	0.13097	0.05777	24	9.96	133.00	1
134	Serine (%)	0504	2.1650	0.13000	2.4185	0.16236	0.06681	26	-1.56	134.00	0
134	Serine (%)	0148	2.2000	0.14000	2.4185	0.16236	0.06681	26	-1.35	134.05	0
134	Serine (%)	0910	2.2350	0.11000	2.4185	0.16236	0.06681	26	-1.13	134.00	0
134	Serine (%)	0098	2.2530	0.09800	2.4185	0.16236	0.06681	26	-1.02	134.02	0
134	Serine (%)	0644	2.2920	0.02200	2.4185	0.16236	0.06681	26	-0.78	134.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
134	Serine (%)	0626	2.3010	0.01000	2.4185	0.16236	0.06681	26	-0.72	134.05	0
134	Serine (%)	0870	2.3023	0.02050	2.4185	0.16236	0.06681	26	-0.72	134.00	0
134	Serine (%)	0859	2.3523	0.01110	2.4185	0.16236	0.06681	26	-0.41	134.00	0
134	Serine (%)	0652	2.3600	0.02000	2.4185	0.16236	0.06681	26	-0.36	134.00	0
134	Serine (%)	0872	2.3665	0.10500	2.4185	0.16236	0.06681	26	-0.32	134.00	0
134	Serine (%)	0684	2.3700	0.02600	2.4185	0.16236	0.06681	26	-0.30	134.00	0
134	Serine (%)	0610	2.3700	0.04000	2.4185	0.16236	0.06681	26	-0.30	134.05	0
134	Serine (%)	0968	2.3720	0.00400	2.4185	0.16236	0.06681	26	-0.29	134.00	0
134	Serine (%)	0571	2.3780	0.02400	2.4185	0.16236	0.06681	26	-0.25	134.00	0
134	Serine (%)	0354	2.3920	0.02000	2.4185	0.16236	0.06681	26	-0.16	134.00	0
134	Serine (%)	2059	2.4305	0.00100	2.4185	0.16236	0.06681	26	0.07	134.00	0
134	Serine (%)	0227	2.4900	0.00000	2.4185	0.16236	0.06681	26	0.44	134.00	0
134	Serine (%)	0878	2.4975	0.17700	2.4185	0.16236	0.06681	26	0.49	134.00	0
134	Serine (%)	0675	2.5050	0.01000	2.4185	0.16236	0.06681	26	0.53	134.00	0
134	Serine (%)	0297	2.5050	0.09600	2.4185	0.16236	0.06681	26	0.53	134.01	0
134	Serine (%)	2144	2.5850	0.09000	2.4185	0.16236	0.06681	26	1.03	134.99	0
134	Serine (%)	0918	2.5955	0.04520	2.4185	0.16236	0.06681	26	1.09	134.00	0
134	Serine (%)	0619	2.6050	0.05000	2.4185	0.16236	0.06681	26	1.15	134.00	0
134	Serine (%)	0676	2.6330	0.05400	2.4185	0.16236	0.06681	26	1.32	134.05	0
134	Serine (%)	0171	2.6545	0.04700	2.4185	0.16236	0.06681	26	1.45	134.00	0
134	Serine (%)	0889	3.0669	0.38630	2.4185	0.16236	0.06681	26	3.99	134.99	0
134	Serine (%)	0226	2.6510	0.47800	2.4185	0.16236	0.06681	26	1.43	134.00	1
135	Threonine (%)	0910	1.7750	0.07000	1.8791	0.08017	0.03491	27	-1.30	135.00	0
135	Threonine (%)	0226	1.7925	0.10700	1.8791	0.08017	0.03491	27	-1.08	135.00	0
135	Threonine (%)	0148	1.8000	0.12000	1.8791	0.08017	0.03491	27	-0.99	135.05	0
135	Threonine (%)	0626	1.8015	0.07500	1.8791	0.08017	0.03491	27	-0.97	135.05	0
135	Threonine (%)	0878	1.8060	0.04200	1.8791	0.08017	0.03491	27	-0.91	135.00	0
135	Threonine (%)	0859	1.8092	0.01450	1.8791	0.08017	0.03491	27	-0.87	135.00	0
135	Threonine (%)	0652	1.8150	0.01000	1.8791	0.08017	0.03491	27	-0.80	135.00	0
135	Threonine (%)	0504	1.8300	0.02000	1.8791	0.08017	0.03491	27	-0.61	135.00	0
135	Threonine (%)	0872	1.8380	0.06400	1.8791	0.08017	0.03491	27	-0.51	135.00	0
135	Threonine (%)	2059	1.8440	0.01000	1.8791	0.08017	0.03491	27	-0.44	135.00	0
135	Threonine (%)	0098	1.8470	0.00600	1.8791	0.08017	0.03491	27	-0.40	135.02	0
135	Threonine (%)	0644	1.8500	0.02200	1.8791	0.08017	0.03491	27	-0.36	135.00	0
135	Threonine (%)	0354	1.8695	0.02700	1.8791	0.08017	0.03491	27	-0.12	135.00	0
135	Threonine (%)	0571	1.8695	0.00500	1.8791	0.08017	0.03491	27	-0.12	135.00	0
135	Threonine (%)	0676	1.8695	0.03500	1.8791	0.08017	0.03491	27	-0.12	135.05	0
135	Threonine (%)	0610	1.8700	0.04000	1.8791	0.08017	0.03491	27	-0.11	135.05	0
135	Threonine (%)	2144	1.8700	0.06000	1.8791	0.08017	0.03491	27	-0.11	135.99	0
135	Threonine (%)	0968	1.8800	0.00000	1.8791	0.08017	0.03491	27	0.01	135.00	0
135	Threonine (%)	0870	1.9235	0.00930	1.8791	0.08017	0.03491	27	0.55	135.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score		
135	Threonine (%)	0297	1.9305	0.07500	1.8791	0.08017	0.03491	27	0.64	135.01	0
135	Threonine (%)	0684	1.9430	0.01000	1.8791	0.08017	0.03491	27	0.80	135.00	0
135	Threonine (%)	0227	1.9500	0.00000	1.8791	0.08017	0.03491	27	0.88	135.00	0
135	Threonine (%)	0675	1.9550	0.01000	1.8791	0.08017	0.03491	27	0.95	135.00	0
135	Threonine (%)	0171	2.0560	0.02800	1.8791	0.08017	0.03491	27	2.21	135.00	0
135	Threonine (%)	0619	2.0850	0.05000	1.8791	0.08017	0.03491	27	2.57	135.00	0
135	Threonine (%)	0889	2.1011	0.01880	1.8791	0.08017	0.03491	27	2.77	135.99	0
135	Threonine (%)	0918	2.1787	0.01400	1.8791	0.08017	0.03491	27	3.74	135.00	0
135	Threonine (%)	0160	1.5361	0.26750	1.8791	0.08017	0.03491	27	-4.28	135.99	1
136	Tryptophan (%)	0889	0.10835	0.00950	0.41466	0.06794	0.01304	18	-4.51	136.99	0
136	Tryptophan (%)	0297	0.29750	0.00300	0.41466	0.06794	0.01304	18	-1.72	136.01	0
136	Tryptophan (%)	0297	0.29750	0.00300	0.41466	0.06794	0.01304	18	-1.72	136.03	0
136	Tryptophan (%)	0504	0.33000	0.02000	0.41466	0.06794	0.01304	18	-1.25	136.99	0
136	Tryptophan (%)	0878	0.39200	0.00400	0.41466	0.06794	0.01304	18	-0.33	136.01	0
136	Tryptophan (%)	0208	0.40450	0.03300	0.41466	0.06794	0.01304	18	-0.15	136.01	0
136	Tryptophan (%)	0910	0.40500	0.03000	0.41466	0.06794	0.01304	18	-0.14	136.00	0
136	Tryptophan (%)	0870	0.41810	0.01460	0.41466	0.06794	0.01304	18	0.05	136.00	0
136	Tryptophan (%)	0684	0.43850	0.00100	0.41466	0.06794	0.01304	18	0.35	136.00	0
136	Tryptophan (%)	0619	0.43900	0.00600	0.41466	0.06794	0.01304	18	0.36	136.00	0
136	Tryptophan (%)	0098	0.44250	0.00100	0.41466	0.06794	0.01304	18	0.41	136.02	0
136	Tryptophan (%)	0859	0.44745	0.02790	0.41466	0.06794	0.01304	18	0.48	136.03	0
136	Tryptophan (%)	2059	0.45500	0.00000	0.41466	0.06794	0.01304	18	0.59	136.03	0
136	Tryptophan (%)	0918	0.45510	0.03880	0.41466	0.06794	0.01304	18	0.60	136.00	0
136	Tryptophan (%)	0872	0.45700	0.00400	0.41466	0.06794	0.01304	18	0.62	136.03	0
136	Tryptophan (%)	0571	0.46750	0.00700	0.41466	0.06794	0.01304	18	0.78	136.03	0
136	Tryptophan (%)	0644	0.46900	0.00200	0.41466	0.06794	0.01304	18	0.80	136.01	0
136	Tryptophan (%)	0227	0.50500	0.03000	0.41466	0.06794	0.01304	18	1.33	136.00	0
137	Tyrosine (%)	0910	0.99500	0.05000	1.4782	0.09257	0.03730	21	-5.22	137.00	0
137	Tyrosine (%)	0889	1.2013	0.00940	1.4782	0.09257	0.03730	21	-2.99	137.99	0
137	Tyrosine (%)	0226	1.3675	0.06900	1.4782	0.09257	0.03730	21	-1.20	137.00	0
137	Tyrosine (%)	0354	1.4040	0.00200	1.4782	0.09257	0.03730	21	-0.80	137.00	0
137	Tyrosine (%)	0870	1.4183	0.00530	1.4782	0.09257	0.03730	21	-0.65	137.00	0
137	Tyrosine (%)	0684	1.4195	0.03100	1.4782	0.09257	0.03730	21	-0.63	137.00	0
137	Tyrosine (%)	0872	1.4340	0.03800	1.4782	0.09257	0.03730	21	-0.48	137.00	0
137	Tyrosine (%)	0968	1.4540	0.06600	1.4782	0.09257	0.03730	21	-0.26	137.00	0
137	Tyrosine (%)	0227	1.4550	0.03000	1.4782	0.09257	0.03730	21	-0.25	137.00	0
137	Tyrosine (%)	0626	1.4605	0.12100	1.4782	0.09257	0.03730	21	-0.19	137.05	0
137	Tyrosine (%)	0644	1.4855	0.00700	1.4782	0.09257	0.03730	21	0.08	137.00	0
137	Tyrosine (%)	0148	1.4950	0.13000	1.4782	0.09257	0.03730	21	0.18	137.05	0
137	Tyrosine (%)	0675	1.5050	0.01000	1.4782	0.09257	0.03730	21	0.29	137.00	0
137	Tyrosine (%)	0098	1.5050	0.00600	1.4782	0.09257	0.03730	21	0.29	137.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
137	Tyrosine (%)	0297	1.5120	0.02200	1.4782	0.09257	0.03730	21	0.36	137.01	0
137	Tyrosine (%)	2059	1.5265	0.01300	1.4782	0.09257	0.03730	21	0.52	137.00	0
137	Tyrosine (%)	0504	1.5350	0.07000	1.4782	0.09257	0.03730	21	0.61	137.00	0
137	Tyrosine (%)	0918	1.5645	0.07270	1.4782	0.09257	0.03730	21	0.93	137.00	0
137	Tyrosine (%)	0676	1.5890	0.00800	1.4782	0.09257	0.03730	21	1.20	137.05	0
137	Tyrosine (%)	0171	1.6265	0.01300	1.4782	0.09257	0.03730	21	1.60	137.00	0
137	Tyrosine (%)	0610	1.7350	0.01000	1.4782	0.09257	0.03730	21	2.77	137.05	0
138	Valine (%)	0889	2.0446	0.01880	2.5187	0.11571	0.04229	26	-4.10	138.99	0
138	Valine (%)	0870	2.3076	0.09200	2.5187	0.11571	0.04229	26	-1.82	138.00	0
138	Valine (%)	0652	2.3700	0.02000	2.5187	0.11571	0.04229	26	-1.29	138.00	0
138	Valine (%)	0226	2.3760	0.19800	2.5187	0.11571	0.04229	26	-1.23	138.00	0
138	Valine (%)	0859	2.4041	0.00550	2.5187	0.11571	0.04229	26	-0.99	138.00	0
138	Valine (%)	0675	2.4050	0.01000	2.5187	0.11571	0.04229	26	-0.98	138.00	0
138	Valine (%)	0626	2.4505	0.03700	2.5187	0.11571	0.04229	26	-0.59	138.05	0
138	Valine (%)	0354	2.4615	0.04900	2.5187	0.11571	0.04229	26	-0.49	138.00	0
138	Valine (%)	0910	2.4800	0.14000	2.5187	0.11571	0.04229	26	-0.33	138.00	0
138	Valine (%)	0878	2.4865	0.03300	2.5187	0.11571	0.04229	26	-0.28	138.00	0
138	Valine (%)	0171	2.5040	0.04600	2.5187	0.11571	0.04229	26	-0.13	138.00	0
138	Valine (%)	0968	2.5245	0.00300	2.5187	0.11571	0.04229	26	0.05	138.00	0
138	Valine (%)	0098	2.5340	0.03400	2.5187	0.11571	0.04229	26	0.13	138.02	0
138	Valine (%)	0676	2.5360	0.03800	2.5187	0.11571	0.04229	26	0.15	138.05	0
138	Valine (%)	2059	2.5395	0.00500	2.5187	0.11571	0.04229	26	0.18	138.00	0
138	Valine (%)	0610	2.5400	0.02000	2.5187	0.11571	0.04229	26	0.18	138.05	0
138	Valine (%)	0684	2.5620	0.03000	2.5187	0.11571	0.04229	26	0.37	138.00	0
138	Valine (%)	0571	2.5650	0.04400	2.5187	0.11571	0.04229	26	0.40	138.00	0
138	Valine (%)	0872	2.5830	0.07000	2.5187	0.11571	0.04229	26	0.56	138.00	0
138	Valine (%)	0148	2.6000	0.00000	2.5187	0.11571	0.04229	26	0.70	138.05	0
138	Valine (%)	0504	2.6050	0.05000	2.5187	0.11571	0.04229	26	0.75	138.00	0
138	Valine (%)	0619	2.6050	0.01000	2.5187	0.11571	0.04229	26	0.75	138.00	0
138	Valine (%)	0644	2.6335	0.00500	2.5187	0.11571	0.04229	26	0.99	138.00	0
138	Valine (%)	0297	2.6465	0.04500	2.5187	0.11571	0.04229	26	1.10	138.01	0
138	Valine (%)	0227	2.7000	0.08000	2.5187	0.11571	0.04229	26	1.57	138.00	0
138	Valine (%)	0918	2.7486	0.01620	2.5187	0.11571	0.04229	26	1.99	138.00	0
138	Valine (%)	2144	2.6250	0.33000	2.5187	0.11571	0.04229	26	0.92	138.99	1
139	Taurine (%)	0644	0.20400	0.00400	0.26581	0.04908	0.00958	6	-1.26	139.00	0
139	Taurine (%)	0889	0.23085	0.00950	0.26581	0.04908	0.00958	6	-0.71	139.99	0
139	Taurine (%)	0098	0.26400	0.00000	0.26581	0.04908	0.00958	6	-0.04	139.02	0
139	Taurine (%)	0171	0.28100	0.01400	0.26581	0.04908	0.00958	6	0.31	139.00	0
139	Taurine (%)	0227	0.29000	0.00000	0.26581	0.04908	0.00958	6	0.49	139.00	0
139	Taurine (%)	0504	0.32500	0.03000	0.26581	0.04908	0.00958	6	1.21	139.00	0
357	Ethoxyquin (mg/kg (ppm))	0227	73.500	1.0000				2		357.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
357	Ethoxyquin (mg/kg (ppm))	0942	74.000	0.00000				2		357.99	0
400	Water activity (Units)	2073	0.35900	0.00000	0.39652	0.03555	0.00448	9	-1.06	400.99	0
400	Water activity (Units)	0589	0.37550	0.00900	0.39652	0.03555	0.00448	9	-0.59	400.01	0
400	Water activity (Units)	0843	0.37900	0.00600	0.39652	0.03555	0.00448	9	-0.49	400.01	0
400	Water activity (Units)	0843	0.38450	0.00100	0.39652	0.03555	0.00448	9	-0.34	400.99	0
400	Water activity (Units)	0407	0.38730	0.00360	0.39652	0.03555	0.00448	9	-0.26	400.01	0
400	Water activity (Units)	0942	0.39100	0.01200	0.39652	0.03555	0.00448	9	-0.16	400.01	0
400	Water activity (Units)	0051	0.39500	0.00400	0.39652	0.03555	0.00448	9	-0.04	400.01	0
400	Water activity (Units)	2009	0.44800	0.00000	0.39652	0.03555	0.00448	9	1.45	400.01	0
400	Water activity (Units)	0552	0.44935	0.00470	0.39652	0.03555	0.00448	9	1.49	400.01	0
412	Dietary Starch (%)	0353	8.6500	0.02000				1		412.01	0
516	Arsenic, total (mg / kg (ppm))	0045	0.41500	0.05000	2.9832	0.26465	0.12190	13	-9.70	516.00	0
516	Arsenic, total (mg / kg (ppm))	2146	2.7330	0.11200	2.9832	0.26465	0.12190	13	-0.95	516.52	0
516	Arsenic, total (mg / kg (ppm))	0171	2.7450	0.01000	2.9832	0.26465	0.12190	13	-0.90	516.00	0
516	Arsenic, total (mg / kg (ppm))	0186	2.8400	0.06000	2.9832	0.26465	0.12190	13	-0.54	516.52	0
516	Arsenic, total (mg / kg (ppm))	0553	2.8850	0.03000	2.9832	0.26465	0.12190	13	-0.37	516.53	0
516	Arsenic, total (mg / kg (ppm))	0227	2.9000	0.12000	2.9832	0.26465	0.12190	13	-0.31	516.53	0
516	Arsenic, total (mg / kg (ppm))	0413	2.9700	0.04000	2.9832	0.26465	0.12190	13	-0.05	516.53	0
516	Arsenic, total (mg / kg (ppm))	0910	2.9750	0.21000	2.9832	0.26465	0.12190	13	-0.03	516.52	0
516	Arsenic, total (mg / kg (ppm))	2012	3.1100	0.04000	2.9832	0.26465	0.12190	13	0.48	516.53	0
516	Arsenic, total (mg / kg (ppm))	0098	3.1910	0.19200	2.9832	0.26465	0.12190	13	0.79	516.53	0
516	Arsenic, total (mg / kg (ppm))	0918	3.2225	0.28990	2.9832	0.26465	0.12190	13	0.90	516.53	0
516	Arsenic, total (mg / kg (ppm))	0508	3.2538	0.41080	2.9832	0.26465	0.12190	13	1.02	516.43	0
516	Arsenic, total (mg / kg (ppm))	2033	3.3700	0.02000	2.9832	0.26465	0.12190	13	1.46	516.43	0
518	Cadmium (mg / kg (ppm))	0045	0.10000	0.00000	0.11034	0.00726	0.00754	12	-1.42	518.42	0
518	Cadmium (mg / kg (ppm))	0227	0.10050	0.01700	0.11034	0.00726	0.00754	12	-1.36	518.53	0
518	Cadmium (mg / kg (ppm))	2012	0.10500	0.01000	0.11034	0.00726	0.00754	12	-0.74	518.53	0
518	Cadmium (mg / kg (ppm))	0918	0.10835	0.00950	0.11034	0.00726	0.00754	12	-0.27	518.53	0
518	Cadmium (mg / kg (ppm))	2033	0.11000	0.00000	0.11034	0.00726	0.00754	12	-0.05	518.43	0
518	Cadmium (mg / kg (ppm))	0186	0.11050	0.00100	0.11034	0.00726	0.00754	12	0.02	518.52	0
518	Cadmium (mg / kg (ppm))	2146	0.11050	0.00500	0.11034	0.00726	0.00754	12	0.02	518.52	0
518	Cadmium (mg / kg (ppm))	0553	0.11200	0.00400	0.11034	0.00726	0.00754	12	0.23	518.53	0
518	Cadmium (mg / kg (ppm))	0171	0.11250	0.01300	0.11034	0.00726	0.00754	12	0.30	518.41	0
518	Cadmium (mg / kg (ppm))	0413	0.11500	0.01000	0.11034	0.00726	0.00754	12	0.64	518.53	0
518	Cadmium (mg / kg (ppm))	0098	0.11850	0.00100	0.11034	0.00726	0.00754	12	1.12	518.53	0
518	Cadmium (mg / kg (ppm))	0910	0.13000	0.02000	0.11034	0.00726	0.00754	12	2.71	518.52	0
520	Chromium (mg / kg (ppm))	0693	2.6240	0.01400	3.1374	0.44192	0.13300	10	-1.16	520.42	0
520	Chromium (mg / kg (ppm))	0186	2.6750	0.19000	3.1374	0.44192	0.13300	10	-1.05	520.52	0
520	Chromium (mg / kg (ppm))	0045	2.9050	0.11000	3.1374	0.44192	0.13300	10	-0.53	520.42	0
520	Chromium (mg / kg (ppm))	0553	2.9600	0.22000	3.1374	0.44192	0.13300	10	-0.40	520.53	0
520	Chromium (mg / kg (ppm))	0171	3.0150	0.11000	3.1374	0.44192	0.13300	10	-0.28	520.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
520	Chromium (mg / kg (ppm))	0510	3.1450	0.15000	3.1374	0.44192	0.13300	10	0.02	520.43	0
520	Chromium (mg / kg (ppm))	2033	3.2000	0.00000	3.1374	0.44192	0.13300	10	0.14	520.43	0
520	Chromium (mg / kg (ppm))	0508	3.3751	0.41260	3.1374	0.44192	0.13300	10	0.54	520.43	0
520	Chromium (mg / kg (ppm))	2053	3.6750	0.05000	3.1374	0.44192	0.13300	10	1.22	520.42	0
520	Chromium (mg / kg (ppm))	0918	4.3021	0.07340	3.1374	0.44192	0.13300	10	2.64	520.53	0
526	Lead (mg / kg (ppm))	0227	0.50000	0.00400	0.56308	0.05115	0.03213	11	-1.23	526.53	0
526	Lead (mg / kg (ppm))	0918	0.52110	0.00040	0.56308	0.05115	0.03213	11	-0.82	526.53	0
526	Lead (mg / kg (ppm))	0553	0.52150	0.00300	0.56308	0.05115	0.03213	11	-0.81	526.53	0
526	Lead (mg / kg (ppm))	0413	0.53000	0.06000	0.56308	0.05115	0.03213	11	-0.65	526.53	0
526	Lead (mg / kg (ppm))	2033	0.53500	0.03000	0.56308	0.05115	0.03213	11	-0.55	526.53	0
526	Lead (mg / kg (ppm))	0098	0.57200	0.04000	0.56308	0.05115	0.03213	11	0.17	526.53	0
526	Lead (mg / kg (ppm))	2146	0.57400	0.03800	0.56308	0.05115	0.03213	11	0.21	526.52	0
526	Lead (mg / kg (ppm))	0186	0.58050	0.00500	0.56308	0.05115	0.03213	11	0.34	526.52	0
526	Lead (mg / kg (ppm))	0910	0.60000	0.06000	0.56308	0.05115	0.03213	11	0.72	526.52	0
526	Lead (mg / kg (ppm))	2012	0.62000	0.06000	0.56308	0.05115	0.03213	11	1.11	526.53	0
526	Lead (mg / kg (ppm))	0171	0.68350	0.05300	0.56308	0.05115	0.03213	11	2.35	526.41	0
529	Mercury (µg / kg (ppb))	0227	16.500	1.0000				2		529.99	0
529	Mercury (µg / kg (ppb))	0918	34.000	2.0000				2		529.99	0
539	Nickel (mg / kg (ppm))	2033	0.84000	0.08000	1.1788	0.37849	0.09066	5		539.43	0
539	Nickel (mg / kg (ppm))	0186	1.0300	0.18000	1.1788	0.37849	0.09066	5		539.52	0
539	Nickel (mg / kg (ppm))	0171	1.0400	0.04000	1.1788	0.37849	0.09066	5		539.41	0
539	Nickel (mg / kg (ppm))	0553	1.1600	0.04000	1.1788	0.37849	0.09066	5		539.53	0
539	Nickel (mg / kg (ppm))	0918	1.8242	0.11330	1.1788	0.37849	0.09066	5		539.53	0
702	Butyric Acid (4:0) (%)	0619	0.02450	0.00300				1		702.00	0
704	Caproic Acid (6:0) (%)	0619	0.00100	0.00000				1		704.00	0
706	Caprylic acid (8:0) (%) (w/w)	0619	0.00000	0.00000				0		706.01	4
708	Capric acid (10:0) (%) (w/w)	0619	0.00300	0.00000				1		708.01	0
710	Lauric Acid (12:0) (%) (w/w)	0910	0.11500	0.01000				2		710.99	0
710	Lauric Acid (12:0) (%) (w/w)	0619	0.14700	0.00200				2		710.01	0
714	Myristic Acid (14:0) (%) (w/w)	0619	1.2700	0.00000				2		714.01	0
714	Myristic Acid (14:0) (%) (w/w)	0910	7.3650	0.15000				2		714.99	0
716	Palmitic Acid (16:0) (%) (w/w)	0619	4.8350	0.05000				2		716.01	0
716	Palmitic Acid (16:0) (%) (w/w)	0910	19.905	0.23000				2		716.99	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0619	2.7150	0.03000				2		718.01	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0910	11.860	0.22000				2		718.99	0
720	Margaric acid (17:0) (%) (w/w)	0910	0.46000	0.00000				1		720.99	0
722	Stearic Acid (18:0) (%) (w/w)	0619	0.69300	0.00800				2		722.01	0
722	Stearic Acid (18:0) (%) (w/w)	0910	3.8150	0.03000				2		722.99	0
724	Oleic Acid (9c-18:1) (%) (w/w)	0619	2.6400	0.04000				2		724.01	0
724	Oleic Acid (9c-18:1) (%) (w/w)	0910	10.715	0.03000				2		724.99	0
726	Linoleic Acid (9c,12c-18:2) (%) (w/w)	0619	1.2850	0.01000				2		726.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0910	5.9500	0.04000				2	726.99	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (%)	0619	0.27000	0.00400				2	728.01	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (%)	0910	1.6450	0.01000				2	728.99	0	
730	Arachidic Acid (20:0) (% (w/w))	0619	0.15100	0.00200				2	730.01	0	
730	Arachidic Acid (20:0) (% (w/w))	0910	0.21500	0.03000				2	730.99	0	
732	Gondoic Acid (11c-20:1) (% (w/w))	0619	0.18150	0.00100				2	732.01	0	
732	Gondoic Acid (11c-20:1) (% (w/w))	0910	0.50000	0.02000				2	732.99	0	
736	Arachidonic Acid (5c,8c,11c,14c-20:4) (%)	0910	1.2700	0.00000				1	736.99	0	
736	Arachidonic Acid (5c,8c,11c,14c-20:4) (%)	0619	0.00000	0.00000				1	736.01	4	
738	Mead Acid (11c,14c,17c-20:3) (% (w/w))	0619	0.16700	0.00400				1	738.01	0	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c-20:5) (%)	0619	1.7200	0.00000				2	740.01	0	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c-20:5) (%)	0910	9.4050	0.09000				2	740.99	0	
744	Erucic Acid (13c-22:1) (% (w/w))	0910	0.08500	0.01000				1	744.99	0	
744	Erucic Acid (13c-22:1) (% (w/w))	0619	0.00000	0.00000				1	744.01	4	
746	Docosapentaenoic Acid n-3 DPA (DHA) (%)	0619	0.29950	0.00300				1	746.01	0	
748	Lignoceric Acid (24:0) (% (w/w))	0910	0.14500	0.01000				1	748.99	0	
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22:6) (%)	0619	0.92700	0.00400				1	750.01	0	
752	Nervonic Acid (24:1) isomers (% (w/w))	0619	0.00000	0.00000				0	752.01	4	

Notes: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2 and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = Rejected for duplicates too far apart, 2 = Rejected as outlier, 5 = A reporting limit and 4 = zeros submitted as values. Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). Flag 3 indicates not used in statistics.