



Equine Feed

Animal Feed Scheme

Labs Reporting: 178

Test Material Code # 201821

Analyte Proficiency Testing Report

Issue Date : 02/28/2018

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
000	Urea (%)	0278	0.10000	0.00000				1		000.02	0
001	Loss on Drying (%)	0618	4.9200	0.10000	7.3017	0.27462	0.08407	71	-8.67	001.07	0
001	Loss on Drying (%)	0618	5.8500	0.34000	7.3017	0.27462	0.08407	71	-5.29	001.99	0
001	Loss on Drying (%)	0536	6.1650	0.17000	7.3017	0.27462	0.08407	71	-4.14	001.99	0
001	Loss on Drying (%)	0074	6.7650	0.07000	7.3017	0.27462	0.08407	71	-1.95	001.07	0
001	Loss on Drying (%)	0676	6.8500	0.02000	7.3017	0.27462	0.08407	71	-1.64	001.99	0
001	Loss on Drying (%)	0019	6.9000	0.20000	7.3017	0.27462	0.08407	71	-1.46	001.07	0
001	Loss on Drying (%)	0630	6.9500	0.00000	7.3017	0.27462	0.08407	71	-1.28	001.99	0
001	Loss on Drying (%)	0918	6.9850	0.21000	7.3017	0.27462	0.08407	71	-1.15	001.99	0
001	Loss on Drying (%)	2126	6.9950	0.01000	7.3017	0.27462	0.08407	71	-1.12	001.99	0
001	Loss on Drying (%)	0089	7.0000	0.00000	7.3017	0.27462	0.08407	71	-1.10	001.07	0
001	Loss on Drying (%)	0612	7.0000	0.04000	7.3017	0.27462	0.08407	71	-1.10	001.99	0
001	Loss on Drying (%)	0889	7.0700	0.00000	7.3017	0.27462	0.08407	71	-0.84	001.99	0
001	Loss on Drying (%)	0948	7.0900	0.04000	7.3017	0.27462	0.08407	71	-0.77	001.99	0
001	Loss on Drying (%)	0407	7.0962	0.08620	7.3017	0.27462	0.08407	71	-0.75	001.07	0
001	Loss on Drying (%)	2195	7.1000	0.00000	7.3017	0.27462	0.08407	71	-0.73	001.07	0
001	Loss on Drying (%)	0910	7.1000	0.00000	7.3017	0.27462	0.08407	71	-0.73	001.99	0
001	Loss on Drying (%)	2155	7.1000	0.20000	7.3017	0.27462	0.08407	71	-0.73	001.99	0
001	Loss on Drying (%)	0297	7.1250	0.19000	7.3017	0.27462	0.08407	71	-0.64	001.07	0
001	Loss on Drying (%)	0413	7.1500	0.10000	7.3017	0.27462	0.08407	71	-0.55	001.07	0
001	Loss on Drying (%)	2146	7.1500	0.04000	7.3017	0.27462	0.08407	71	-0.55	001.07	0
001	Loss on Drying (%)	0083	7.1550	0.03000	7.3017	0.27462	0.08407	71	-0.53	001.07	0
001	Loss on Drying (%)	2059	7.1600	0.02000	7.3017	0.27462	0.08407	71	-0.52	001.07	0
001	Loss on Drying (%)	2009	7.1624	0.01040	7.3017	0.27462	0.08407	71	-0.51	001.07	0
001	Loss on Drying (%)	0610	7.1650	0.01000	7.3017	0.27462	0.08407	71	-0.50	001.05	0
001	Loss on Drying (%)	0504	7.1700	0.08000	7.3017	0.27462	0.08407	71	-0.48	001.00	0
001	Loss on Drying (%)	0034	7.1900	0.24000	7.3017	0.27462	0.08407	71	-0.41	001.99	0
001	Loss on Drying (%)	2053	7.2100	0.06000	7.3017	0.27462	0.08407	71	-0.33	001.07	0
001	Loss on Drying (%)	2145	7.2100	0.00000	7.3017	0.27462	0.08407	71	-0.33	001.07	0
001	Loss on Drying (%)	0139	7.2150	0.01000	7.3017	0.27462	0.08407	71	-0.32	001.07	0
001	Loss on Drying (%)	0675	7.2250	0.07000	7.3017	0.27462	0.08407	71	-0.28	001.07	0
001	Loss on Drying (%)	0016	7.2300	0.18000	7.3017	0.27462	0.08407	71	-0.26	001.00	0
001	Loss on Drying (%)	0309	7.2480	0.01400	7.3017	0.27462	0.08407	71	-0.20	001.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
001	Loss on Drying (%)	0353	7.2500	0.02000	7.3017	0.27462	0.08407	71	-0.19	001.07	0
001	Loss on Drying (%)	0049	7.2600	0.10000	7.3017	0.27462	0.08407	71	-0.15	001.07	0
001	Loss on Drying (%)	0186	7.2700	0.36000	7.3017	0.27462	0.08407	71	-0.12	001.03	0
001	Loss on Drying (%)	0045	7.2700	0.06000	7.3017	0.27462	0.08407	71	-0.12	001.07	0
001	Loss on Drying (%)	0187	7.3150	0.01000	7.3017	0.27462	0.08407	71	0.05	001.07	0
001	Loss on Drying (%)	0895	7.3250	0.01000	7.3017	0.27462	0.08407	71	0.08	001.03	0
001	Loss on Drying (%)	0003	7.3400	0.12000	7.3017	0.27462	0.08407	71	0.14	001.07	0
001	Loss on Drying (%)	0560	7.3400	0.44000	7.3017	0.27462	0.08407	71	0.14	001.99	0
001	Loss on Drying (%)	2152	7.3500	0.02000	7.3017	0.27462	0.08407	71	0.18	001.03	0
001	Loss on Drying (%)	0510	7.3500	0.10000	7.3017	0.27462	0.08407	71	0.18	001.99	0
001	Loss on Drying (%)	0571	7.3550	0.01000	7.3017	0.27462	0.08407	71	0.19	001.07	0
001	Loss on Drying (%)	0638	7.3600	0.00000	7.3017	0.27462	0.08407	71	0.21	001.07	0
001	Loss on Drying (%)	0868	7.3800	0.00000	7.3017	0.27462	0.08407	71	0.29	001.03	0
001	Loss on Drying (%)	0693	7.4000	0.20000	7.3017	0.27462	0.08407	71	0.36	001.07	0
001	Loss on Drying (%)	0505	7.4100	0.08000	7.3017	0.27462	0.08407	71	0.39	001.99	0
001	Loss on Drying (%)	0035	7.4150	0.01000	7.3017	0.27462	0.08407	71	0.41	001.07	0
001	Loss on Drying (%)	0226	7.4300	0.18000	7.3017	0.27462	0.08407	71	0.47	001.07	0
001	Loss on Drying (%)	0590	7.4400	0.22000	7.3017	0.27462	0.08407	71	0.50	001.07	0
001	Loss on Drying (%)	0171	7.4650	0.01000	7.3017	0.27462	0.08407	71	0.59	001.07	0
001	Loss on Drying (%)	0644	7.4850	0.05000	7.3017	0.27462	0.08407	71	0.67	001.05	0
001	Loss on Drying (%)	0027	7.4940	0.01000	7.3017	0.27462	0.08407	71	0.70	001.00	0
001	Loss on Drying (%)	0629	7.5150	0.09000	7.3017	0.27462	0.08407	71	0.78	001.99	0
001	Loss on Drying (%)	0169	7.5150	0.03000	7.3017	0.27462	0.08407	71	0.78	001.00	0
001	Loss on Drying (%)	0878	7.5200	0.06000	7.3017	0.27462	0.08407	71	0.80	001.03	0
001	Loss on Drying (%)	0581	7.5250	0.03000	7.3017	0.27462	0.08407	71	0.81	001.07	0
001	Loss on Drying (%)	2109	7.5250	0.01000	7.3017	0.27462	0.08407	71	0.81	001.07	0
001	Loss on Drying (%)	0689	7.5500	0.10000	7.3017	0.27462	0.08407	71	0.90	001.07	0
001	Loss on Drying (%)	0849	7.5660	0.03200	7.3017	0.27462	0.08407	71	0.96	001.07	0
001	Loss on Drying (%)	0098	7.5700	0.04000	7.3017	0.27462	0.08407	71	0.98	001.07	0
001	Loss on Drying (%)	0876	7.5750	0.03000	7.3017	0.27462	0.08407	71	1.00	001.07	0
001	Loss on Drying (%)	0142	7.6350	0.05000	7.3017	0.27462	0.08407	71	1.21	001.07	0
001	Loss on Drying (%)	0683	7.6600	0.22000	7.3017	0.27462	0.08407	71	1.30	001.07	0
001	Loss on Drying (%)	0872	7.6950	0.03000	7.3017	0.27462	0.08407	71	1.43	001.07	0
001	Loss on Drying (%)	0037	7.6950	0.05000	7.3017	0.27462	0.08407	71	1.43	001.99	0
001	Loss on Drying (%)	0357	7.8444	0.03660	7.3017	0.27462	0.08407	71	1.98	001.99	0
001	Loss on Drying (%)	0019	7.9400	0.02000	7.3017	0.27462	0.08407	71	2.32	001.99	0
001	Loss on Drying (%)	2192	8.0500	0.20000	7.3017	0.27462	0.08407	71	2.72	001.07	0
001	Loss on Drying (%)	0278	8.4150	0.19000	7.3017	0.27462	0.08407	71	4.05	001.07	0
001	Loss on Drying (%)	0609	8.5000	0.20000	7.3017	0.27462	0.08407	71	4.36	001.07	0
001	Loss on Drying (%)	0512	6.8970	0.51000	7.3017	0.27462	0.08407	71	-1.47	001.07	1
001	Loss on Drying (%)	0015	7.0500	0.50000	7.3017	0.27462	0.08407	71	-0.92	001.07	1

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein (%)	0536	15.715	0.09000	18.545	0.32459	0.14752	173	-8.72	002.03	0
002	Protein (%)	0536	16.525	0.17000	18.545	0.32459	0.14752	173	-6.22	002.05	0
002	Protein (%)	0169	17.770	0.02000	18.545	0.32459	0.14752	173	-2.39	002.00	0
002	Protein (%)	2146	17.775	0.21000	18.545	0.32459	0.14752	173	-2.37	002.05	0
002	Protein (%)	2188	17.815	0.01000	18.545	0.32459	0.14752	173	-2.25	002.01	0
002	Protein (%)	0874	17.820	0.10000	18.545	0.32459	0.14752	173	-2.23	002.04	0
002	Protein (%)	0676	17.970	0.18000	18.545	0.32459	0.14752	173	-1.77	002.06	0
002	Protein (%)	2145	18.025	0.05000	18.545	0.32459	0.14752	173	-1.60	002.05	0
002	Protein (%)	0889	18.040	0.04000	18.545	0.32459	0.14752	173	-1.56	002.01	0
002	Protein (%)	0539	18.075	0.03000	18.545	0.32459	0.14752	173	-1.45	002.06	0
002	Protein (%)	0910	18.100	0.00000	18.545	0.32459	0.14752	173	-1.37	002.01	0
002	Protein (%)	0921	18.110	0.00000	18.545	0.32459	0.14752	173	-1.34	002.05	0
002	Protein (%)	0918	18.135	0.21000	18.545	0.32459	0.14752	173	-1.26	002.06	0
002	Protein (%)	0043	18.160	0.08000	18.545	0.32459	0.14752	173	-1.19	002.01	0
002	Protein (%)	0881	18.171	0.01300	18.545	0.32459	0.14752	173	-1.15	002.05	0
002	Protein (%)	2073	18.175	0.05000	18.545	0.32459	0.14752	173	-1.14	002.05	0
002	Protein (%)	0868	18.180	0.12000	18.545	0.32459	0.14752	173	-1.12	002.05	0
002	Protein (%)	0098	18.185	0.17000	18.545	0.32459	0.14752	173	-1.11	002.06	0
002	Protein (%)	0653	18.190	0.06000	18.545	0.32459	0.14752	173	-1.09	002.01	0
002	Protein (%)	0504	18.190	0.52000	18.545	0.32459	0.14752	173	-1.09	002.04	0
002	Protein (%)	0354	18.195	0.01000	18.545	0.32459	0.14752	173	-1.08	002.05	0
002	Protein (%)	0689	18.200	0.00000	18.545	0.32459	0.14752	173	-1.06	002.05	0
002	Protein (%)	0100	18.200	0.06000	18.545	0.32459	0.14752	173	-1.06	002.06	0
002	Protein (%)	2181	18.215	0.21000	18.545	0.32459	0.14752	173	-1.02	002.05	0
002	Protein (%)	0675	18.220	0.06000	18.545	0.32459	0.14752	173	-1.00	002.05	0
002	Protein (%)	2195	18.240	0.00000	18.545	0.32459	0.14752	173	-0.94	002.01	0
002	Protein (%)	0955	18.250	0.30000	18.545	0.32459	0.14752	173	-0.91	002.05	0
002	Protein (%)	0160	18.250	0.70000	18.545	0.32459	0.14752	173	-0.91	002.06	0
002	Protein (%)	0941	18.260	0.30000	18.545	0.32459	0.14752	173	-0.88	002.05	0
002	Protein (%)	0859	18.269	0.11100	18.545	0.32459	0.14752	173	-0.85	002.06	0
002	Protein (%)	2022	18.275	0.01000	18.545	0.32459	0.14752	173	-0.83	002.05	0
002	Protein (%)	0968	18.276	0.23600	18.545	0.32459	0.14752	173	-0.83	002.01	0
002	Protein (%)	0872	18.280	0.04000	18.545	0.32459	0.14752	173	-0.82	002.06	0
002	Protein (%)	0885	18.284	0.27370	18.545	0.32459	0.14752	173	-0.80	002.05	0
002	Protein (%)	0939	18.285	0.09000	18.545	0.32459	0.14752	173	-0.80	002.05	0
002	Protein (%)	0018	18.285	0.11000	18.545	0.32459	0.14752	173	-0.80	002.06	0
002	Protein (%)	0723	18.290	0.00000	18.545	0.32459	0.14752	173	-0.79	002.01	0
002	Protein (%)	0884	18.300	0.00000	18.545	0.32459	0.14752	173	-0.75	002.06	0
002	Protein (%)	0644	18.308	0.16600	18.545	0.32459	0.14752	173	-0.73	002.06	0
002	Protein (%)	2109	18.310	0.10000	18.545	0.32459	0.14752	173	-0.72	002.05	0
002	Protein (%)	0019	18.315	0.19000	18.545	0.32459	0.14752	173	-0.71	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein (%)	0895	18.320	0.02000	18.545	0.32459	0.14752	173	-0.69	002.05	0
002	Protein (%)	0309	18.337	0.21100	18.545	0.32459	0.14752	173	-0.64	002.06	0
002	Protein (%)	2196	18.340	0.00000	18.545	0.32459	0.14752	173	-0.63	002.05	0
002	Protein (%)	0609	18.340	0.10000	18.545	0.32459	0.14752	173	-0.63	002.06	0
002	Protein (%)	0870	18.342	0.80100	18.545	0.32459	0.14752	173	-0.63	002.01	0
002	Protein (%)	2155	18.343	0.62900	18.545	0.32459	0.14752	173	-0.62	002.06	0
002	Protein (%)	0505	18.345	0.03000	18.545	0.32459	0.14752	173	-0.62	002.06	0
002	Protein (%)	0010	18.350	0.30000	18.545	0.32459	0.14752	173	-0.60	002.06	0
002	Protein (%)	0202	18.350	0.06000	18.545	0.32459	0.14752	173	-0.60	002.06	0
002	Protein (%)	0955	18.350	0.10000	18.545	0.32459	0.14752	173	-0.60	002.06	0
002	Protein (%)	0870	18.357	0.05190	18.545	0.32459	0.14752	173	-0.58	002.06	0
002	Protein (%)	0939	18.360	0.16000	18.545	0.32459	0.14752	173	-0.57	002.06	0
002	Protein (%)	0263	18.364	0.02070	18.545	0.32459	0.14752	173	-0.56	002.06	0
002	Protein (%)	0878	18.370	0.10000	18.545	0.32459	0.14752	173	-0.54	002.05	0
002	Protein (%)	0036	18.373	0.22000	18.545	0.32459	0.14752	173	-0.53	002.02	0
002	Protein (%)	0123	18.380	0.10000	18.545	0.32459	0.14752	173	-0.51	002.06	0
002	Protein (%)	0638	18.390	0.06000	18.545	0.32459	0.14752	173	-0.48	002.04	0
002	Protein (%)	0015	18.395	0.43000	18.545	0.32459	0.14752	173	-0.46	002.06	0
002	Protein (%)	0592	18.395	0.07000	18.545	0.32459	0.14752	173	-0.46	002.06	0
002	Protein (%)	0598	18.395	0.15000	18.545	0.32459	0.14752	173	-0.46	002.06	0
002	Protein (%)	0014	18.400	0.20000	18.545	0.32459	0.14752	173	-0.45	002.06	0
002	Protein (%)	0226	18.405	0.39000	18.545	0.32459	0.14752	173	-0.43	002.06	0
002	Protein (%)	0009	18.415	0.15000	18.545	0.32459	0.14752	173	-0.40	002.06	0
002	Protein (%)	2168	18.415	0.07000	18.545	0.32459	0.14752	173	-0.40	002.06	0
002	Protein (%)	2150	18.417	0.25800	18.545	0.32459	0.14752	173	-0.39	002.06	0
002	Protein (%)	2009	18.425	0.08030	18.545	0.32459	0.14752	173	-0.37	002.05	0
002	Protein (%)	2006	18.425	0.09000	18.545	0.32459	0.14752	173	-0.37	002.05	0
002	Protein (%)	0144	18.425	0.01000	18.545	0.32459	0.14752	173	-0.37	002.06	0
002	Protein (%)	0098	18.430	0.08000	18.545	0.32459	0.14752	173	-0.35	002.08	0
002	Protein (%)	0520	18.435	0.15000	18.545	0.32459	0.14752	173	-0.34	002.06	0
002	Protein (%)	2178	18.448	0.45300	18.545	0.32459	0.14752	173	-0.30	002.06	0
002	Protein (%)	0956	18.450	0.50000	18.545	0.32459	0.14752	173	-0.29	002.06	0
002	Protein (%)	0652	18.450	0.30000	18.545	0.32459	0.14752	173	-0.29	002.01	0
002	Protein (%)	0868	18.455	0.05000	18.545	0.32459	0.14752	173	-0.28	002.06	0
002	Protein (%)	0948	18.475	0.03000	18.545	0.32459	0.14752	173	-0.22	002.05	0
002	Protein (%)	0042	18.480	0.16000	18.545	0.32459	0.14752	173	-0.20	002.02	0
002	Protein (%)	2152	18.480	0.02000	18.545	0.32459	0.14752	173	-0.20	002.05	0
002	Protein (%)	0139	18.485	0.03000	18.545	0.32459	0.14752	173	-0.18	002.06	0
002	Protein (%)	0848	18.490	0.02000	18.545	0.32459	0.14752	173	-0.17	002.06	0
002	Protein (%)	0948	18.490	0.10000	18.545	0.32459	0.14752	173	-0.17	002.06	0
002	Protein (%)	0510	18.500	0.00000	18.545	0.32459	0.14752	173	-0.14	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein (%)	0626	18.500	0.02000	18.545	0.32459	0.14752	173	-0.14	002.06	0
002	Protein (%)	0897	18.505	0.03000	18.545	0.32459	0.14752	173	-0.12	002.05	0
002	Protein (%)	0171	18.510	0.04000	18.545	0.32459	0.14752	173	-0.11	002.06	0
002	Protein (%)	0297	18.520	0.02000	18.545	0.32459	0.14752	173	-0.08	002.06	0
002	Protein (%)	0693	18.524	0.08100	18.545	0.32459	0.14752	173	-0.07	002.06	0
002	Protein (%)	0187	18.535	0.01000	18.545	0.32459	0.14752	173	-0.03	002.04	0
002	Protein (%)	0857	18.535	0.21000	18.545	0.32459	0.14752	173	-0.03	002.06	0
002	Protein (%)	0004	18.540	0.02000	18.545	0.32459	0.14752	173	-0.02	002.06	0
002	Protein (%)	0027	18.540	0.12000	18.545	0.32459	0.14752	173	-0.02	002.06	0
002	Protein (%)	0683	18.545	0.19000	18.545	0.32459	0.14752	173	0.00	002.05	0
002	Protein (%)	0294	18.550	0.10000	18.545	0.32459	0.14752	173	0.02	002.06	0
002	Protein (%)	0726	18.550	0.10000	18.545	0.32459	0.14752	173	0.02	002.06	0
002	Protein (%)	0512	18.560	0.12000	18.545	0.32459	0.14752	173	0.05	002.06	0
002	Protein (%)	0675	18.565	0.01000	18.545	0.32459	0.14752	173	0.06	002.06	0
002	Protein (%)	0964	18.574	0.02500	18.545	0.32459	0.14752	173	0.09	002.06	0
002	Protein (%)	0610	18.575	0.05000	18.545	0.32459	0.14752	173	0.09	002.06	0
002	Protein (%)	0504	18.580	0.04000	18.545	0.32459	0.14752	173	0.11	002.06	0
002	Protein (%)	0242	18.595	0.15000	18.545	0.32459	0.14752	173	0.15	002.06	0
002	Protein (%)	0032	18.596	0.19400	18.545	0.32459	0.14752	173	0.16	002.06	0
002	Protein (%)	0849	18.597	0.06700	18.545	0.32459	0.14752	173	0.16	002.05	0
002	Protein (%)	0205	18.600	0.04000	18.545	0.32459	0.14752	173	0.17	002.06	0
002	Protein (%)	0278	18.600	0.20000	18.545	0.32459	0.14752	173	0.17	002.06	0
002	Protein (%)	0345	18.600	0.20000	18.545	0.32459	0.14752	173	0.17	002.06	0
002	Protein (%)	0354	18.600	0.06000	18.545	0.32459	0.14752	173	0.17	002.06	0
002	Protein (%)	0034	18.609	0.18800	18.545	0.32459	0.14752	173	0.20	002.06	0
002	Protein (%)	0074	18.620	0.02000	18.545	0.32459	0.14752	173	0.23	002.06	0
002	Protein (%)	0646	18.625	0.11000	18.545	0.32459	0.14752	173	0.25	002.06	0
002	Protein (%)	0571	18.636	0.13400	18.545	0.32459	0.14752	173	0.28	002.06	0
002	Protein (%)	0629	18.640	0.12000	18.545	0.32459	0.14752	173	0.29	002.10	0
002	Protein (%)	0175	18.650	0.10000	18.545	0.32459	0.14752	173	0.32	002.06	0
002	Protein (%)	0413	18.650	0.10000	18.545	0.32459	0.14752	173	0.32	002.06	0
002	Protein (%)	0682	18.650	0.00000	18.545	0.32459	0.14752	173	0.32	002.06	0
002	Protein (%)	0028	18.655	0.35000	18.545	0.32459	0.14752	173	0.34	002.01	0
002	Protein (%)	0425	18.655	0.11000	18.545	0.32459	0.14752	173	0.34	002.06	0
002	Protein (%)	0089	18.660	0.00000	18.545	0.32459	0.14752	173	0.35	002.06	0
002	Protein (%)	2126	18.660	0.02000	18.545	0.32459	0.14752	173	0.35	002.06	0
002	Protein (%)	0358	18.665	0.05000	18.545	0.32459	0.14752	173	0.37	002.06	0
002	Protein (%)	0142	18.670	0.00000	18.545	0.32459	0.14752	173	0.39	002.06	0
002	Protein (%)	0590	18.675	0.15000	18.545	0.32459	0.14752	173	0.40	002.06	0
002	Protein (%)	0003	18.680	0.16000	18.545	0.32459	0.14752	173	0.42	002.06	0
002	Protein (%)	0687	18.680	0.06000	18.545	0.32459	0.14752	173	0.42	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein (%)	0190	18.685	0.09000	18.545	0.32459	0.14752	173	0.43	002.06	0
002	Protein (%)	0043	18.690	0.38000	18.545	0.32459	0.14752	173	0.45	002.06	0
002	Protein (%)	0045	18.700	0.20000	18.545	0.32459	0.14752	173	0.48	002.06	0
002	Protein (%)	0108	18.705	0.17000	18.545	0.32459	0.14752	173	0.49	002.06	0
002	Protein (%)	0049	18.725	0.05000	18.545	0.32459	0.14752	173	0.55	002.06	0
002	Protein (%)	0618	18.738	0.06300	18.545	0.32459	0.14752	173	0.59	002.06	0
002	Protein (%)	0037	18.740	0.10000	18.545	0.32459	0.14752	173	0.60	002.06	0
002	Protein (%)	0035	18.750	0.02000	18.545	0.32459	0.14752	173	0.63	002.06	0
002	Protein (%)	2193	18.770	0.22000	18.545	0.32459	0.14752	173	0.69	002.02	0
002	Protein (%)	2089	18.790	0.10000	18.545	0.32459	0.14752	173	0.75	002.06	0
002	Protein (%)	0589	18.795	0.03000	18.545	0.32459	0.14752	173	0.77	002.06	0
002	Protein (%)	0407	18.811	0.06000	18.545	0.32459	0.14752	173	0.82	002.06	0
002	Protein (%)	0357	18.813	0.00000	18.545	0.32459	0.14752	173	0.82	002.06	0
002	Protein (%)	0042	18.815	0.25000	18.545	0.32459	0.14752	173	0.83	002.06	0
002	Protein (%)	0300	18.820	0.34000	18.545	0.32459	0.14752	173	0.85	002.06	0
002	Protein (%)	0353	18.835	0.11000	18.545	0.32459	0.14752	173	0.89	002.06	0
002	Protein (%)	0968	18.840	0.20000	18.545	0.32459	0.14752	173	0.91	002.06	0
002	Protein (%)	0910	18.850	0.10000	18.545	0.32459	0.14752	173	0.94	002.06	0
002	Protein (%)	0643	18.850	0.48000	18.545	0.32459	0.14752	173	0.94	002.99	0
002	Protein (%)	0152	18.861	0.15500	18.545	0.32459	0.14752	173	0.97	002.01	0
002	Protein (%)	0889	18.880	0.38000	18.545	0.32459	0.14752	173	1.03	002.11	0
002	Protein (%)	0229	18.890	0.14000	18.545	0.32459	0.14752	173	1.06	002.06	0
002	Protein (%)	2059	18.908	0.05310	18.545	0.32459	0.14752	173	1.12	002.06	0
002	Protein (%)	0508	18.915	0.20000	18.545	0.32459	0.14752	173	1.14	002.06	0
002	Protein (%)	0036	18.935	0.03000	18.545	0.32459	0.14752	173	1.20	002.06	0
002	Protein (%)	0685	18.950	0.18000	18.545	0.32459	0.14752	173	1.25	002.01	0
002	Protein (%)	2076	18.957	0.08300	18.545	0.32459	0.14752	173	1.27	002.06	0
002	Protein (%)	0233	18.965	0.01000	18.545	0.32459	0.14752	173	1.29	002.06	0
002	Protein (%)	2006	18.972	0.23700	18.545	0.32459	0.14752	173	1.31	002.06	0
002	Protein (%)	0630	18.980	0.52000	18.545	0.32459	0.14752	173	1.34	002.06	0
002	Protein (%)	0186	19.000	0.20000	18.545	0.32459	0.14752	173	1.40	002.06	0
002	Protein (%)	0083	19.015	0.15000	18.545	0.32459	0.14752	173	1.45	002.06	0
002	Protein (%)	0683	19.040	0.12000	18.545	0.32459	0.14752	173	1.52	002.06	0
002	Protein (%)	0001	19.045	0.27000	18.545	0.32459	0.14752	173	1.54	002.06	0
002	Protein (%)	0574	19.095	0.05000	18.545	0.32459	0.14752	173	1.69	002.06	0
002	Protein (%)	0208	19.100	0.60000	18.545	0.32459	0.14752	173	1.71	002.06	0
002	Protein (%)	2192	19.125	0.55000	18.545	0.32459	0.14752	173	1.79	002.11	0
002	Protein (%)	0062	19.130	0.10700	18.545	0.32459	0.14752	173	1.80	002.06	0
002	Protein (%)	0511	19.140	0.12000	18.545	0.32459	0.14752	173	1.83	002.06	0
002	Protein (%)	2129	19.200	0.00000	18.545	0.32459	0.14752	173	2.02	002.06	0
002	Protein (%)	0541	19.340	0.68000	18.545	0.32459	0.14752	173	2.45	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein (%)	0011	19.420	0.00000	18.545	0.32459	0.14752	173	2.70	002.11	0
002	Protein (%)	2081	19.495	0.49000	18.545	0.32459	0.14752	173	2.93	002.06	0
002	Protein (%)	2105	19.875	0.01000	18.545	0.32459	0.14752	173	4.10	002.11	0
002	Protein (%)	0953	20.105	0.15000	18.545	0.32459	0.14752	173	4.81	002.06	0
002	Protein (%)	0297	20.430	0.02000	18.545	0.32459	0.14752	173	5.81	002.11	0
002	Protein (%)	0612	20.888	0.60000	18.545	0.32459	0.14752	173	7.22	002.06	0
002	Protein (%)	0876	21.110	0.28000	18.545	0.32459	0.14752	173	7.90	002.06	0
002	Protein (%)	2120	21.975	0.25000	18.545	0.32459	0.14752	173	10.57	002.11	0
002	Protein (%)	2121	22.135	0.31000	18.545	0.32459	0.14752	173	11.06	002.11	0
002	Protein (%)	0692	18.500	1.2000	18.545	0.32459	0.14752	173	-0.14	002.06	1
002	Protein (%)	0011	19.170	1.6000	18.545	0.32459	0.14752	173	1.93	002.06	1
002	Protein (%)	0952	13.645	0.03000	18.545	0.32459	0.14752	173	-15.10	002.05	2
002	Protein (%)	0405	23.570	0.08000	18.545	0.32459	0.14752	173	15.48	002.04	2
003	Fat (%)	2022	2.1700	0.00000	3.2956	0.75924	0.15207	133	-1.48	003.10	0
003	Fat (%)	2155	2.1980	0.41200	3.2956	0.75924	0.15207	133	-1.45	003.10	0
003	Fat (%)	0229	2.2030	0.03000	3.2956	0.75924	0.15207	133	-1.44	003.14	0
003	Fat (%)	0229	2.2130	0.00000	3.2956	0.75924	0.15207	133	-1.43	003.06	0
003	Fat (%)	0263	2.2415	0.03700	3.2956	0.75924	0.15207	133	-1.39	003.09	0
003	Fat (%)	0142	2.3100	0.16000	3.2956	0.75924	0.15207	133	-1.30	003.00	0
003	Fat (%)	0027	2.3195	0.14900	3.2956	0.75924	0.15207	133	-1.29	003.09	0
003	Fat (%)	0952	2.3300	0.02000	3.2956	0.75924	0.15207	133	-1.27	003.10	0
003	Fat (%)	0885	2.3732	0.05730	3.2956	0.75924	0.15207	133	-1.21	003.00	0
003	Fat (%)	0358	2.3850	0.19000	3.2956	0.75924	0.15207	133	-1.20	003.14	0
003	Fat (%)	0045	2.4100	0.18000	3.2956	0.75924	0.15207	133	-1.17	003.10	0
003	Fat (%)	0019	2.4150	0.01000	3.2956	0.75924	0.15207	133	-1.16	003.14	0
003	Fat (%)	0848	2.4950	0.17000	3.2956	0.75924	0.15207	133	-1.05	003.14	0
003	Fat (%)	0089	2.5000	0.00000	3.2956	0.75924	0.15207	133	-1.05	003.14	0
003	Fat (%)	0144	2.5000	0.28000	3.2956	0.75924	0.15207	133	-1.05	003.14	0
003	Fat (%)	2195	2.5300	0.00000	3.2956	0.75924	0.15207	133	-1.01	003.00	0
003	Fat (%)	0897	2.5450	0.01000	3.2956	0.75924	0.15207	133	-0.99	003.10	0
003	Fat (%)	0004	2.5500	0.14000	3.2956	0.75924	0.15207	133	-0.98	003.14	0
003	Fat (%)	0175	2.5500	0.02000	3.2956	0.75924	0.15207	133	-0.98	003.14	0
003	Fat (%)	2053	2.5550	0.09000	3.2956	0.75924	0.15207	133	-0.98	003.00	0
003	Fat (%)	0629	2.5700	0.10000	3.2956	0.75924	0.15207	133	-0.96	003.10	0
003	Fat (%)	0357	2.6000	0.20000	3.2956	0.75924	0.15207	133	-0.92	003.14	0
003	Fat (%)	0870	2.6375	0.05010	3.2956	0.75924	0.15207	133	-0.87	003.10	0
003	Fat (%)	0171	2.6450	0.07000	3.2956	0.75924	0.15207	133	-0.86	003.12	0
003	Fat (%)	0638	2.6500	0.00000	3.2956	0.75924	0.15207	133	-0.85	003.09	0
003	Fat (%)	0108	2.6500	0.42000	3.2956	0.75924	0.15207	133	-0.85	003.14	0
003	Fat (%)	0242	2.6500	0.18000	3.2956	0.75924	0.15207	133	-0.85	003.10	0
003	Fat (%)	2009	2.6510	0.03280	3.2956	0.75924	0.15207	133	-0.85	003.10	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat (%)	0001	2.6710	0.06800	3.2956	0.75924	0.15207	133	-0.82	003.14	0
003	Fat (%)	0034	2.7100	0.02000	3.2956	0.75924	0.15207	133	-0.77	003.14	0
003	Fat (%)	0233	2.7200	0.08000	3.2956	0.75924	0.15207	133	-0.76	003.10	0
003	Fat (%)	0512	2.7355	0.39700	3.2956	0.75924	0.15207	133	-0.74	003.09	0
003	Fat (%)	0037	2.7450	0.51000	3.2956	0.75924	0.15207	133	-0.73	003.14	0
003	Fat (%)	0202	2.7450	0.05000	3.2956	0.75924	0.15207	133	-0.73	003.14	0
003	Fat (%)	0049	2.7500	0.84000	3.2956	0.75924	0.15207	133	-0.72	003.14	0
003	Fat (%)	0009	2.7550	0.01000	3.2956	0.75924	0.15207	133	-0.71	003.14	0
003	Fat (%)	0035	2.7800	0.08000	3.2956	0.75924	0.15207	133	-0.68	003.00	0
003	Fat (%)	0062	2.7880	0.16200	3.2956	0.75924	0.15207	133	-0.67	003.10	0
003	Fat (%)	2193	2.7900	0.28000	3.2956	0.75924	0.15207	133	-0.67	003.99	0
003	Fat (%)	0881	2.8295	0.11300	3.2956	0.75924	0.15207	133	-0.61	003.10	0
003	Fat (%)	0868	2.8300	0.02000	3.2956	0.75924	0.15207	133	-0.61	003.10	0
003	Fat (%)	0123	2.8350	0.05000	3.2956	0.75924	0.15207	133	-0.61	003.13	0
003	Fat (%)	0692	2.8500	0.10000	3.2956	0.75924	0.15207	133	-0.59	003.14	0
003	Fat (%)	0004	2.8700	0.28000	3.2956	0.75924	0.15207	133	-0.56	003.09	0
003	Fat (%)	2129	2.8900	0.24000	3.2956	0.75924	0.15207	133	-0.53	003.14	0
003	Fat (%)	0100	2.9100	0.32000	3.2956	0.75924	0.15207	133	-0.51	003.10	0
003	Fat (%)	0968	2.9270	0.00400	3.2956	0.75924	0.15207	133	-0.49	003.14	0
003	Fat (%)	2120	2.9300	0.66000	3.2956	0.75924	0.15207	133	-0.48	003.11	0
003	Fat (%)	0876	3.0000	0.18000	3.2956	0.75924	0.15207	133	-0.39	003.14	0
003	Fat (%)	0098	3.0350	0.03000	3.2956	0.75924	0.15207	133	-0.34	003.10	0
003	Fat (%)	0948	3.0400	0.10000	3.2956	0.75924	0.15207	133	-0.34	003.14	0
003	Fat (%)	0675	3.0500	0.02000	3.2956	0.75924	0.15207	133	-0.32	003.09	0
003	Fat (%)	0520	3.0650	0.67000	3.2956	0.75924	0.15207	133	-0.30	003.14	0
003	Fat (%)	0036	3.0786	0.38290	3.2956	0.75924	0.15207	133	-0.29	003.10	0
003	Fat (%)	2196	3.1000	0.00000	3.2956	0.75924	0.15207	133	-0.26	003.13	0
003	Fat (%)	0015	3.1150	0.17000	3.2956	0.75924	0.15207	133	-0.24	003.14	0
003	Fat (%)	0889	3.1200	0.00000	3.2956	0.75924	0.15207	133	-0.23	003.10	0
003	Fat (%)	0309	3.1264	0.15810	3.2956	0.75924	0.15207	133	-0.22	003.00	0
003	Fat (%)	0186	3.1300	0.04000	3.2956	0.75924	0.15207	133	-0.22	003.14	0
003	Fat (%)	0407	3.1389	0.08400	3.2956	0.75924	0.15207	133	-0.21	003.14	0
003	Fat (%)	2146	3.1550	0.09000	3.2956	0.75924	0.15207	133	-0.19	003.14	0
003	Fat (%)	0682	3.1600	0.00000	3.2956	0.75924	0.15207	133	-0.18	003.14	0
003	Fat (%)	0878	3.1600	0.10000	3.2956	0.75924	0.15207	133	-0.18	003.14	0
003	Fat (%)	0689	3.2000	0.20000	3.2956	0.75924	0.15207	133	-0.13	003.06	0
003	Fat (%)	0646	3.2050	0.15000	3.2956	0.75924	0.15207	133	-0.12	003.13	0
003	Fat (%)	2188	3.2200	0.02000	3.2956	0.75924	0.15207	133	-0.10	003.10	0
003	Fat (%)	2192	3.2300	0.04000	3.2956	0.75924	0.15207	133	-0.09	003.00	0
003	Fat (%)	2073	3.2350	0.31000	3.2956	0.75924	0.15207	133	-0.08	003.10	0
003	Fat (%)	0939	3.2400	0.14000	3.2956	0.75924	0.15207	133	-0.07	003.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat (%)	0689	3.2500	0.10000	3.2956	0.75924	0.15207	133	-0.06	003.10	0
003	Fat (%)	0187	3.2550	0.11000	3.2956	0.75924	0.15207	133	-0.05	003.13	0
003	Fat (%)	0098	3.2700	0.38000	3.2956	0.75924	0.15207	133	-0.03	003.13	0
003	Fat (%)	0941	3.2750	0.13000	3.2956	0.75924	0.15207	133	-0.03	003.99	0
003	Fat (%)	0294	3.2800	0.00000	3.2956	0.75924	0.15207	133	-0.02	003.06	0
003	Fat (%)	0152	3.3000	0.00000	3.2956	0.75924	0.15207	133	0.01	003.00	0
003	Fat (%)	0205	3.3050	0.25000	3.2956	0.75924	0.15207	133	0.01	003.13	0
003	Fat (%)	0505	3.3300	0.10000	3.2956	0.75924	0.15207	133	0.05	003.09	0
003	Fat (%)	0074	3.3450	0.03000	3.2956	0.75924	0.15207	133	0.07	003.14	0
003	Fat (%)	0849	3.3505	0.02100	3.2956	0.75924	0.15207	133	0.07	003.06	0
003	Fat (%)	2121	3.3900	0.80000	3.2956	0.75924	0.15207	133	0.12	003.11	0
003	Fat (%)	0598	3.3950	0.33000	3.2956	0.75924	0.15207	133	0.13	003.14	0
003	Fat (%)	0226	3.4200	0.10000	3.2956	0.75924	0.15207	133	0.16	003.09	0
003	Fat (%)	0955	3.4200	0.14000	3.2956	0.75924	0.15207	133	0.16	003.14	0
003	Fat (%)	0683	3.4250	0.05000	3.2956	0.75924	0.15207	133	0.17	003.14	0
003	Fat (%)	0098	3.4300	0.04000	3.2956	0.75924	0.15207	133	0.18	003.09	0
003	Fat (%)	0723	3.4300	0.00000	3.2956	0.75924	0.15207	133	0.18	003.09	0
003	Fat (%)	2076	3.4694	0.54020	3.2956	0.75924	0.15207	133	0.23	003.14	0
003	Fat (%)	0581	3.4700	0.08000	3.2956	0.75924	0.15207	133	0.23	003.14	0
003	Fat (%)	0536	3.4850	0.03000	3.2956	0.75924	0.15207	133	0.25	003.99	0
003	Fat (%)	0175	3.5100	0.06000	3.2956	0.75924	0.15207	133	0.28	003.00	0
003	Fat (%)	0028	3.5100	0.08000	3.2956	0.75924	0.15207	133	0.28	003.12	0
003	Fat (%)	0278	3.5500	0.10000	3.2956	0.75924	0.15207	133	0.34	003.14	0
003	Fat (%)	0956	3.5500	0.50000	3.2956	0.75924	0.15207	133	0.34	003.14	0
003	Fat (%)	0508	3.5584	0.23940	3.2956	0.75924	0.15207	133	0.35	003.09	0
003	Fat (%)	0047	3.6600	0.14000	3.2956	0.75924	0.15207	133	0.48	003.99	0
003	Fat (%)	0693	3.6770	0.14800	3.2956	0.75924	0.15207	133	0.50	003.10	0
003	Fat (%)	0644	3.6940	0.16600	3.2956	0.75924	0.15207	133	0.52	003.06	0
003	Fat (%)	0190	3.6950	0.05000	3.2956	0.75924	0.15207	133	0.53	003.14	0
003	Fat (%)	0413	3.7000	0.40000	3.2956	0.75924	0.15207	133	0.53	003.14	0
003	Fat (%)	2181	3.8000	0.04000	3.2956	0.75924	0.15207	133	0.66	003.06	0
003	Fat (%)	0011	3.8500	0.00000	3.2956	0.75924	0.15207	133	0.73	003.11	0
003	Fat (%)	0160	3.8500	0.52000	3.2956	0.75924	0.15207	133	0.73	003.99	0
003	Fat (%)	0574	3.8850	0.03000	3.2956	0.75924	0.15207	133	0.78	003.14	0
003	Fat (%)	0726	3.8900	0.18000	3.2956	0.75924	0.15207	133	0.78	003.14	0
003	Fat (%)	0609	3.9450	0.07000	3.2956	0.75924	0.15207	133	0.86	003.10	0
003	Fat (%)	0618	4.0330	0.01400	3.2956	0.75924	0.15207	133	0.97	003.10	0
003	Fat (%)	0504	4.0650	0.21000	3.2956	0.75924	0.15207	133	1.01	003.99	0
003	Fat (%)	0345	4.0750	0.05000	3.2956	0.75924	0.15207	133	1.03	003.00	0
003	Fat (%)	0353	4.0950	0.47000	3.2956	0.75924	0.15207	133	1.05	003.09	0
003	Fat (%)	0884	4.1000	0.00000	3.2956	0.75924	0.15207	133	1.06	003.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat (%)	0511	4.1200	0.68000	3.2956	0.75924	0.15207	133	1.09	003.06	0
003	Fat (%)	0354	4.2050	0.03000	3.2956	0.75924	0.15207	133	1.20	003.09	0
003	Fat (%)	0676	4.2100	0.00000	3.2956	0.75924	0.15207	133	1.20	003.10	0
003	Fat (%)	2109	4.2450	0.13000	3.2956	0.75924	0.15207	133	1.25	003.00	0
003	Fat (%)	0297	4.3200	0.08000	3.2956	0.75924	0.15207	133	1.35	003.06	0
003	Fat (%)	2006	4.3300	0.30000	3.2956	0.75924	0.15207	133	1.36	003.13	0
003	Fat (%)	0948	4.3350	0.05000	3.2956	0.75924	0.15207	133	1.37	003.09	0
003	Fat (%)	0918	4.3550	0.01000	3.2956	0.75924	0.15207	133	1.40	003.06	0
003	Fat (%)	0889	4.4150	0.07000	3.2956	0.75924	0.15207	133	1.47	003.11	0
003	Fat (%)	0685	4.4550	0.05000	3.2956	0.75924	0.15207	133	1.53	003.09	0
003	Fat (%)	0895	4.4600	0.00000	3.2956	0.75924	0.15207	133	1.53	003.12	0
003	Fat (%)	0653	4.5000	0.02000	3.2956	0.75924	0.15207	133	1.59	003.00	0
003	Fat (%)	0541	4.5600	0.18000	3.2956	0.75924	0.15207	133	1.67	003.10	0
003	Fat (%)	0910	4.5700	0.46000	3.2956	0.75924	0.15207	133	1.68	003.06	0
003	Fat (%)	2089	4.6250	0.05000	3.2956	0.75924	0.15207	133	1.75	003.10	0
003	Fat (%)	2126	4.6450	0.05000	3.2956	0.75924	0.15207	133	1.78	003.00	0
003	Fat (%)	0590	4.6650	0.17000	3.2956	0.75924	0.15207	133	1.80	003.09	0
003	Fat (%)	0510	4.7000	0.00000	3.2956	0.75924	0.15207	133	1.85	003.09	0
003	Fat (%)	0630	4.8350	0.11000	3.2956	0.75924	0.15207	133	2.03	003.09	0
003	Fat (%)	0682	4.9800	0.00000	3.2956	0.75924	0.15207	133	2.22	003.06	0
003	Fat (%)	0032	5.0000	0.20000	3.2956	0.75924	0.15207	133	2.24	003.06	0
003	Fat (%)	0083	5.0450	0.63000	3.2956	0.75924	0.15207	133	2.30	003.06	0
003	Fat (%)	2105	5.3000	0.08000	3.2956	0.75924	0.15207	133	2.64	003.11	0
003	Fat (%)	0003	3.3950	1.3900	3.2956	0.75924	0.15207	133	0.13	003.14	1
003	Fat (%)	0208	3.7000	1.5800	3.2956	0.75924	0.15207	133	0.53	003.13	1
004	Fiber (%)	2089	8.4550	0.11000	11.147	0.86775	0.21605	105	-3.10	004.03	0
004	Fiber (%)	2121	8.6950	0.01000	11.147	0.86775	0.21605	105	-2.83	004.11	0
004	Fiber (%)	2120	9.0850	0.01000	11.147	0.86775	0.21605	105	-2.38	004.11	0
004	Fiber (%)	0921	9.6600	0.00000	11.147	0.86775	0.21605	105	-1.71	004.03	0
004	Fiber (%)	0682	9.8000	0.00000	11.147	0.86775	0.21605	105	-1.55	004.07	0
004	Fiber (%)	0910	9.8500	0.30000	11.147	0.86775	0.21605	105	-1.49	004.07	0
004	Fiber (%)	0425	9.8500	0.10000	11.147	0.86775	0.21605	105	-1.49	004.00	0
004	Fiber (%)	2181	9.8850	0.13000	11.147	0.86775	0.21605	105	-1.45	004.07	0
004	Fiber (%)	0581	10.100	0.60000	11.147	0.86775	0.21605	105	-1.21	004.07	0
004	Fiber (%)	0505	10.105	0.07000	11.147	0.86775	0.21605	105	-1.20	004.07	0
004	Fiber (%)	0171	10.110	0.16000	11.147	0.86775	0.21605	105	-1.19	004.00	0
004	Fiber (%)	2146	10.135	0.39000	11.147	0.86775	0.21605	105	-1.17	004.06	0
004	Fiber (%)	0123	10.150	0.10000	11.147	0.86775	0.21605	105	-1.15	004.07	0
004	Fiber (%)	0683	10.180	0.04000	11.147	0.86775	0.21605	105	-1.11	004.07	0
004	Fiber (%)	0011	10.250	0.00000	11.147	0.86775	0.21605	105	-1.03	004.11	0
004	Fiber (%)	0918	10.270	0.12000	11.147	0.86775	0.21605	105	-1.01	004.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber (%)	0692	10.300	0.20000	11.147	0.86775	0.21605	105	-0.98	004.07	0
004	Fiber (%)	0035	10.345	0.67000	11.147	0.86775	0.21605	105	-0.92	004.07	0
004	Fiber (%)	0675	10.355	0.03000	11.147	0.86775	0.21605	105	-0.91	004.06	0
004	Fiber (%)	0100	10.365	0.91000	11.147	0.86775	0.21605	105	-0.90	004.07	0
004	Fiber (%)	0895	10.415	0.01000	11.147	0.86775	0.21605	105	-0.84	004.06	0
004	Fiber (%)	0345	10.500	0.20000	11.147	0.86775	0.21605	105	-0.75	004.00	0
004	Fiber (%)	2073	10.515	0.07000	11.147	0.86775	0.21605	105	-0.73	004.00	0
004	Fiber (%)	0202	10.520	0.16000	11.147	0.86775	0.21605	105	-0.72	004.07	0
004	Fiber (%)	2145	10.565	0.11000	11.147	0.86775	0.21605	105	-0.67	004.06	0
004	Fiber (%)	0723	10.590	0.00000	11.147	0.86775	0.21605	105	-0.64	004.06	0
004	Fiber (%)	0646	10.620	0.02000	11.147	0.86775	0.21605	105	-0.61	004.07	0
004	Fiber (%)	0003	10.660	0.90000	11.147	0.86775	0.21605	105	-0.56	004.07	0
004	Fiber (%)	0726	10.750	0.10000	11.147	0.86775	0.21605	105	-0.46	004.07	0
004	Fiber (%)	0956	10.750	0.50000	11.147	0.86775	0.21605	105	-0.46	004.07	0
004	Fiber (%)	0629	10.750	0.00000	11.147	0.86775	0.21605	105	-0.46	004.99	0
004	Fiber (%)	0644	10.791	0.13300	11.147	0.86775	0.21605	105	-0.41	004.07	0
004	Fiber (%)	0028	10.800	0.40000	11.147	0.86775	0.21605	105	-0.40	004.07	0
004	Fiber (%)	0190	10.800	0.00000	11.147	0.86775	0.21605	105	-0.40	004.07	0
004	Fiber (%)	0897	10.805	0.01000	11.147	0.86775	0.21605	105	-0.39	004.06	0
004	Fiber (%)	0098	10.825	0.31000	11.147	0.86775	0.21605	105	-0.37	004.07	0
004	Fiber (%)	0884	10.850	0.30000	11.147	0.86775	0.21605	105	-0.34	004.00	0
004	Fiber (%)	0708	10.850	0.52000	11.147	0.86775	0.21605	105	-0.34	004.07	0
004	Fiber (%)	0889	10.865	0.33000	11.147	0.86775	0.21605	105	-0.32	004.11	0
004	Fiber (%)	0870	10.891	0.27770	11.147	0.86775	0.21605	105	-0.29	004.07	0
004	Fiber (%)	2022	10.895	0.03000	11.147	0.86775	0.21605	105	-0.29	004.06	0
004	Fiber (%)	0511	10.900	0.60000	11.147	0.86775	0.21605	105	-0.28	004.00	0
004	Fiber (%)	2053	10.930	0.02000	11.147	0.86775	0.21605	105	-0.25	004.07	0
004	Fiber (%)	0169	10.935	0.17000	11.147	0.86775	0.21605	105	-0.24	004.00	0
004	Fiber (%)	0032	10.950	0.16000	11.147	0.86775	0.21605	105	-0.23	004.07	0
004	Fiber (%)	0848	10.950	0.00000	11.147	0.86775	0.21605	105	-0.23	004.07	0
004	Fiber (%)	0590	10.970	0.38000	11.147	0.86775	0.21605	105	-0.20	004.06	0
004	Fiber (%)	0885	10.979	0.09600	11.147	0.86775	0.21605	105	-0.19	004.06	0
004	Fiber (%)	0939	11.005	0.21000	11.147	0.86775	0.21605	105	-0.16	004.07	0
004	Fiber (%)	0205	11.050	0.10000	11.147	0.86775	0.21605	105	-0.11	004.06	0
004	Fiber (%)	0175	11.075	0.15000	11.147	0.86775	0.21605	105	-0.08	004.00	0
004	Fiber (%)	0353	11.085	0.07000	11.147	0.86775	0.21605	105	-0.07	004.03	0
004	Fiber (%)	0354	11.115	0.03000	11.147	0.86775	0.21605	105	-0.04	004.07	0
004	Fiber (%)	2188	11.125	0.15000	11.147	0.86775	0.21605	105	-0.02	004.07	0
004	Fiber (%)	0354	11.135	0.01000	11.147	0.86775	0.21605	105	-0.01	004.00	0
004	Fiber (%)	0098	11.140	1.1600	11.147	0.86775	0.21605	105	-0.01	004.06	0
004	Fiber (%)	2006	11.140	0.14000	11.147	0.86775	0.21605	105	-0.01	004.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber (%)	0278	11.150	0.50000	11.147	0.86775	0.21605	105	0.00	004.07	0
004	Fiber (%)	0226	11.160	0.46000	11.147	0.86775	0.21605	105	0.02	004.00	0
004	Fiber (%)	2076	11.180	0.03390	11.147	0.86775	0.21605	105	0.04	004.00	0
004	Fiber (%)	0263	11.187	0.10000	11.147	0.86775	0.21605	105	0.05	004.07	0
004	Fiber (%)	2152	11.190	0.02000	11.147	0.86775	0.21605	105	0.05	004.06	0
004	Fiber (%)	0941	11.240	0.12000	11.147	0.86775	0.21605	105	0.11	004.06	0
004	Fiber (%)	0229	11.245	0.03000	11.147	0.86775	0.21605	105	0.11	004.07	0
004	Fiber (%)	0413	11.350	0.70000	11.147	0.86775	0.21605	105	0.23	004.07	0
004	Fiber (%)	0510	11.350	0.10000	11.147	0.86775	0.21605	105	0.23	004.07	0
004	Fiber (%)	0689	11.400	0.00000	11.147	0.86775	0.21605	105	0.29	004.06	0
004	Fiber (%)	2109	11.400	0.18000	11.147	0.86775	0.21605	105	0.29	004.07	0
004	Fiber (%)	0868	11.415	0.01000	11.147	0.86775	0.21605	105	0.31	004.06	0
004	Fiber (%)	0074	11.420	0.06000	11.147	0.86775	0.21605	105	0.32	004.07	0
004	Fiber (%)	0643	11.445	0.51000	11.147	0.86775	0.21605	105	0.34	004.07	0
004	Fiber (%)	0309	11.458	0.14500	11.147	0.86775	0.21605	105	0.36	004.00	0
004	Fiber (%)	0208	11.500	0.20000	11.147	0.86775	0.21605	105	0.41	004.00	0
004	Fiber (%)	0242	11.510	0.34000	11.147	0.86775	0.21605	105	0.42	004.07	0
004	Fiber (%)	0881	11.525	0.58000	11.147	0.86775	0.21605	105	0.44	004.06	0
004	Fiber (%)	0638	11.550	0.10000	11.147	0.86775	0.21605	105	0.46	004.06	0
004	Fiber (%)	0294	11.550	0.10000	11.147	0.86775	0.21605	105	0.46	004.07	0
004	Fiber (%)	0878	11.555	0.17000	11.147	0.86775	0.21605	105	0.47	004.06	0
004	Fiber (%)	2192	11.560	0.32000	11.147	0.86775	0.21605	105	0.48	004.03	0
004	Fiber (%)	0948	11.570	0.06000	11.147	0.86775	0.21605	105	0.49	004.06	0
004	Fiber (%)	0876	11.645	0.51000	11.147	0.86775	0.21605	105	0.57	004.07	0
004	Fiber (%)	0512	11.685	0.13000	11.147	0.86775	0.21605	105	0.62	004.06	0
004	Fiber (%)	0849	11.698	0.18500	11.147	0.86775	0.21605	105	0.63	004.07	0
004	Fiber (%)	0952	11.855	0.01000	11.147	0.86775	0.21605	105	0.82	004.06	0
004	Fiber (%)	0948	11.995	0.09000	11.147	0.86775	0.21605	105	0.98	004.07	0
004	Fiber (%)	2009	12.026	0.22950	11.147	0.86775	0.21605	105	1.01	004.07	0
004	Fiber (%)	0968	12.034	0.05400	11.147	0.86775	0.21605	105	1.02	004.06	0
004	Fiber (%)	0089	12.090	0.08000	11.147	0.86775	0.21605	105	1.09	004.07	0
004	Fiber (%)	0009	12.120	0.04000	11.147	0.86775	0.21605	105	1.12	004.07	0
004	Fiber (%)	0504	12.190	0.20000	11.147	0.86775	0.21605	105	1.20	004.00	0
004	Fiber (%)	0955	12.300	0.00000	11.147	0.86775	0.21605	105	1.33	004.07	0
004	Fiber (%)	0653	12.320	0.08000	11.147	0.86775	0.21605	105	1.35	004.07	0
004	Fiber (%)	0968	12.346	0.48900	11.147	0.86775	0.21605	105	1.38	004.07	0
004	Fiber (%)	2146	12.355	0.75000	11.147	0.86775	0.21605	105	1.39	004.07	0
004	Fiber (%)	0019	12.395	0.03000	11.147	0.86775	0.21605	105	1.44	004.07	0
004	Fiber (%)	0144	12.430	0.42000	11.147	0.86775	0.21605	105	1.48	004.07	0
004	Fiber (%)	2196	12.580	0.00000	11.147	0.86775	0.21605	105	1.65	004.07	0
004	Fiber (%)	0034	12.605	1.0300	11.147	0.86775	0.21605	105	1.68	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber (%)	0045	12.650	0.30000	11.147	0.86775	0.21605	105	1.73	004.03	0
004	Fiber (%)	0407	12.728	0.38370	11.147	0.86775	0.21605	105	1.82	004.07	0
004	Fiber (%)	0609	12.900	0.12000	11.147	0.86775	0.21605	105	2.02	004.06	0
004	Fiber (%)	0520	13.250	0.10000	11.147	0.86775	0.21605	105	2.42	004.07	0
004	Fiber (%)	0693	13.255	0.24800	11.147	0.86775	0.21605	105	2.43	004.03	0
004	Fiber (%)	0598	13.605	0.39000	11.147	0.86775	0.21605	105	2.83	004.07	0
004	Fiber (%)	0536	14.840	0.20000	11.147	0.86775	0.21605	105	4.26	004.07	0
004	Fiber (%)	0160	9.2900	2.8200	11.147	0.86775	0.21605	105	-2.14	004.07	1
004	Fiber (%)	2195	17.500	2.2000	11.147	0.86775	0.21605	105	7.32	004.07	1
005	Ash (%)	0609	10.815	0.07000	11.950	0.39325	0.10936	125	-2.89	005.00	0
005	Ash (%)	0169	10.980	0.04000	11.950	0.39325	0.10936	125	-2.47	005.00	0
005	Ash (%)	2181	11.045	0.03000	11.950	0.39325	0.10936	125	-2.30	005.00	0
005	Ash (%)	0539	11.070	0.08000	11.950	0.39325	0.10936	125	-2.24	005.00	0
005	Ash (%)	0049	11.185	0.23000	11.950	0.39325	0.10936	125	-1.94	005.00	0
005	Ash (%)	0278	11.205	0.03000	11.950	0.39325	0.10936	125	-1.89	005.00	0
005	Ash (%)	0015	11.275	0.03000	11.950	0.39325	0.10936	125	-1.72	005.00	0
005	Ash (%)	0541	11.275	0.01000	11.950	0.39325	0.10936	125	-1.72	005.00	0
005	Ash (%)	0175	11.300	0.16000	11.950	0.39325	0.10936	125	-1.65	005.00	0
005	Ash (%)	0511	11.300	0.00000	11.950	0.39325	0.10936	125	-1.65	005.00	0
005	Ash (%)	2022	11.370	0.08000	11.950	0.39325	0.10936	125	-1.47	005.00	0
005	Ash (%)	0027	11.380	0.00500	11.950	0.39325	0.10936	125	-1.45	005.00	0
005	Ash (%)	0208	11.400	0.60000	11.950	0.39325	0.10936	125	-1.40	005.00	0
005	Ash (%)	0956	11.400	0.00000	11.950	0.39325	0.10936	125	-1.40	005.00	0
005	Ash (%)	0520	11.405	0.05000	11.950	0.39325	0.10936	125	-1.38	005.00	0
005	Ash (%)	0205	11.420	0.02000	11.950	0.39325	0.10936	125	-1.35	005.00	0
005	Ash (%)	0598	11.430	0.12000	11.950	0.39325	0.10936	125	-1.32	005.00	0
005	Ash (%)	0881	11.446	0.06700	11.950	0.39325	0.10936	125	-1.28	005.00	0
005	Ash (%)	0357	11.450	0.30000	11.950	0.39325	0.10936	125	-1.27	005.00	0
005	Ash (%)	0638	11.490	0.18000	11.950	0.39325	0.10936	125	-1.17	005.05	0
005	Ash (%)	0003	11.500	0.00000	11.950	0.39325	0.10936	125	-1.14	005.05	0
005	Ash (%)	0004	11.510	0.04000	11.950	0.39325	0.10936	125	-1.12	005.00	0
005	Ash (%)	0574	11.515	0.05000	11.950	0.39325	0.10936	125	-1.11	005.99	0
005	Ash (%)	0726	11.545	0.21000	11.950	0.39325	0.10936	125	-1.03	005.05	0
005	Ash (%)	0160	11.550	0.30000	11.950	0.39325	0.10936	125	-1.02	005.00	0
005	Ash (%)	0897	11.575	0.03000	11.950	0.39325	0.10936	125	-0.95	005.05	0
005	Ash (%)	0723	11.594	0.00000	11.950	0.39325	0.10936	125	-0.91	005.00	0
005	Ash (%)	2073	11.600	0.14000	11.950	0.39325	0.10936	125	-0.89	005.00	0
005	Ash (%)	0512	11.630	0.32000	11.950	0.39325	0.10936	125	-0.81	005.00	0
005	Ash (%)	0693	11.677	0.55400	11.950	0.39325	0.10936	125	-0.69	005.00	0
005	Ash (%)	0358	11.690	0.12000	11.950	0.39325	0.10936	125	-0.66	005.00	0
005	Ash (%)	0019	11.705	0.61000	11.950	0.39325	0.10936	125	-0.62	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0186	11.750	0.10000	11.950	0.39325	0.10936	125	-0.51	005.00	0
005	Ash (%)	0876	11.760	0.02000	11.950	0.39325	0.10936	125	-0.48	005.00	0
005	Ash (%)	0889	11.775	0.09000	11.950	0.39325	0.10936	125	-0.44	005.11	0
005	Ash (%)	0653	11.790	0.02000	11.950	0.39325	0.10936	125	-0.41	005.00	0
005	Ash (%)	2105	11.790	0.08000	11.950	0.39325	0.10936	125	-0.41	005.11	0
005	Ash (%)	0510	11.805	0.01000	11.950	0.39325	0.10936	125	-0.37	005.00	0
005	Ash (%)	0589	11.815	0.25000	11.950	0.39325	0.10936	125	-0.34	005.00	0
005	Ash (%)	0953	11.830	0.00000	11.950	0.39325	0.10936	125	-0.30	005.00	0
005	Ash (%)	0868	11.835	0.01000	11.950	0.39325	0.10936	125	-0.29	005.05	0
005	Ash (%)	0144	11.840	0.18000	11.950	0.39325	0.10936	125	-0.28	005.00	0
005	Ash (%)	0870	11.854	0.27610	11.950	0.39325	0.10936	125	-0.24	005.00	0
005	Ash (%)	2192	11.860	0.34000	11.950	0.39325	0.10936	125	-0.23	005.05	0
005	Ash (%)	0874	11.865	0.09000	11.950	0.39325	0.10936	125	-0.22	005.00	0
005	Ash (%)	0108	11.870	0.22000	11.950	0.39325	0.10936	125	-0.20	005.00	0
005	Ash (%)	2195	11.890	0.12000	11.950	0.39325	0.10936	125	-0.15	005.00	0
005	Ash (%)	2193	11.895	0.03000	11.950	0.39325	0.10936	125	-0.14	005.00	0
005	Ash (%)	0142	11.920	0.00000	11.950	0.39325	0.10936	125	-0.08	005.00	0
005	Ash (%)	0229	11.920	0.26000	11.950	0.39325	0.10936	125	-0.08	005.00	0
005	Ash (%)	2150	11.920	0.16000	11.950	0.39325	0.10936	125	-0.08	005.00	0
005	Ash (%)	0001	11.923	0.11660	11.950	0.39325	0.10936	125	-0.07	005.00	0
005	Ash (%)	0918	11.925	0.17000	11.950	0.39325	0.10936	125	-0.06	005.00	0
005	Ash (%)	0643	11.930	0.04000	11.950	0.39325	0.10936	125	-0.05	005.00	0
005	Ash (%)	2053	11.930	0.16000	11.950	0.39325	0.10936	125	-0.05	005.00	0
005	Ash (%)	0646	11.935	0.31000	11.950	0.39325	0.10936	125	-0.04	005.00	0
005	Ash (%)	2006	11.935	0.03000	11.950	0.39325	0.10936	125	-0.04	005.00	0
005	Ash (%)	0682	11.940	0.00000	11.950	0.39325	0.10936	125	-0.02	005.00	0
005	Ash (%)	0948	11.940	0.02000	11.950	0.39325	0.10936	125	-0.02	005.00	0
005	Ash (%)	0629	11.940	0.08000	11.950	0.39325	0.10936	125	-0.02	005.99	0
005	Ash (%)	2188	11.955	0.03000	11.950	0.39325	0.10936	125	0.01	005.00	0
005	Ash (%)	0618	11.957	0.14800	11.950	0.39325	0.10936	125	0.02	005.00	0
005	Ash (%)	0849	11.968	0.01600	11.950	0.39325	0.10936	125	0.05	005.05	0
005	Ash (%)	0035	11.995	0.01000	11.950	0.39325	0.10936	125	0.12	005.00	0
005	Ash (%)	0226	12.000	0.46000	11.950	0.39325	0.10936	125	0.13	005.00	0
005	Ash (%)	0263	12.013	0.04000	11.950	0.39325	0.10936	125	0.16	005.05	0
005	Ash (%)	2076	12.013	0.15150	11.950	0.39325	0.10936	125	0.16	005.00	0
005	Ash (%)	0968	12.019	0.00700	11.950	0.39325	0.10936	125	0.18	005.00	0
005	Ash (%)	0187	12.030	0.00000	11.950	0.39325	0.10936	125	0.20	005.00	0
005	Ash (%)	0297	12.035	0.13000	11.950	0.39325	0.10936	125	0.22	005.00	0
005	Ash (%)	2196	12.050	0.00000	11.950	0.39325	0.10936	125	0.26	005.00	0
005	Ash (%)	0689	12.065	0.07000	11.950	0.39325	0.10936	125	0.29	005.05	0
005	Ash (%)	0848	12.065	0.03000	11.950	0.39325	0.10936	125	0.29	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0354	12.075	0.03000	11.950	0.39325	0.10936	125	0.32	005.00	0
005	Ash (%)	0407	12.077	0.24900	11.950	0.39325	0.10936	125	0.32	005.00	0
005	Ash (%)	0504	12.080	0.24000	11.950	0.39325	0.10936	125	0.33	005.00	0
005	Ash (%)	0045	12.100	0.20000	11.950	0.39325	0.10936	125	0.38	005.00	0
005	Ash (%)	0505	12.100	0.04000	11.950	0.39325	0.10936	125	0.38	005.00	0
005	Ash (%)	0294	12.100	0.00000	11.950	0.39325	0.10936	125	0.38	005.05	0
005	Ash (%)	0878	12.100	0.12000	11.950	0.39325	0.10936	125	0.38	005.05	0
005	Ash (%)	0630	12.115	0.01000	11.950	0.39325	0.10936	125	0.42	005.00	0
005	Ash (%)	0098	12.115	0.07000	11.950	0.39325	0.10936	125	0.42	005.00	0
005	Ash (%)	0425	12.120	0.12000	11.950	0.39325	0.10936	125	0.43	005.00	0
005	Ash (%)	0676	12.140	0.02000	11.950	0.39325	0.10936	125	0.48	005.99	0
005	Ash (%)	2009	12.147	0.04670	11.950	0.39325	0.10936	125	0.50	005.05	0
005	Ash (%)	0171	12.150	0.06000	11.950	0.39325	0.10936	125	0.51	005.00	0
005	Ash (%)	0345	12.150	0.10000	11.950	0.39325	0.10936	125	0.51	005.00	0
005	Ash (%)	0083	12.150	0.12000	11.950	0.39325	0.10936	125	0.51	005.00	0
005	Ash (%)	0353	12.150	0.04000	11.950	0.39325	0.10936	125	0.51	005.00	0
005	Ash (%)	0202	12.165	0.05000	11.950	0.39325	0.10936	125	0.55	005.99	0
005	Ash (%)	0683	12.170	0.02000	11.950	0.39325	0.10936	125	0.56	005.05	0
005	Ash (%)	0914	12.175	0.13000	11.950	0.39325	0.10936	125	0.57	005.99	0
005	Ash (%)	0309	12.195	0.07000	11.950	0.39325	0.10936	125	0.62	005.00	0
005	Ash (%)	2089	12.200	0.12000	11.950	0.39325	0.10936	125	0.64	005.00	0
005	Ash (%)	0413	12.200	0.20000	11.950	0.39325	0.10936	125	0.64	005.05	0
005	Ash (%)	0895	12.205	0.01000	11.950	0.39325	0.10936	125	0.65	005.05	0
005	Ash (%)	0590	12.210	0.06000	11.950	0.39325	0.10936	125	0.66	005.05	0
005	Ash (%)	0939	12.230	0.06000	11.950	0.39325	0.10936	125	0.71	005.05	0
005	Ash (%)	0889	12.235	0.03000	11.950	0.39325	0.10936	125	0.73	005.05	0
005	Ash (%)	0885	12.244	0.15280	11.950	0.39325	0.10936	125	0.75	005.05	0
005	Ash (%)	0692	12.250	0.10000	11.950	0.39325	0.10936	125	0.76	005.05	0
005	Ash (%)	0139	12.260	0.02000	11.950	0.39325	0.10936	125	0.79	005.00	0
005	Ash (%)	0536	12.290	0.02000	11.950	0.39325	0.10936	125	0.87	005.99	0
005	Ash (%)	0610	12.300	0.06000	11.950	0.39325	0.10936	125	0.89	005.05	0
005	Ash (%)	2109	12.300	0.00000	11.950	0.39325	0.10936	125	0.89	005.05	0
005	Ash (%)	0884	12.300	0.00000	11.950	0.39325	0.10936	125	0.89	005.99	0
005	Ash (%)	0581	12.315	0.11000	11.950	0.39325	0.10936	125	0.93	005.00	0
005	Ash (%)	0675	12.320	0.08000	11.950	0.39325	0.10936	125	0.94	005.00	0
005	Ash (%)	2146	12.355	0.01000	11.950	0.39325	0.10936	125	1.03	005.05	0
005	Ash (%)	2145	12.360	0.20000	11.950	0.39325	0.10936	125	1.04	005.05	0
005	Ash (%)	0152	12.370	0.22000	11.950	0.39325	0.10936	125	1.07	005.05	0
005	Ash (%)	2126	12.380	0.02000	11.950	0.39325	0.10936	125	1.09	005.05	0
005	Ash (%)	0644	12.395	0.17000	11.950	0.39325	0.10936	125	1.13	005.02	0
005	Ash (%)	0652	12.400	0.00000	11.950	0.39325	0.10936	125	1.15	005.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0190	12.425	0.03000	11.950	0.39325	0.10936	125	1.21	005.05	0
005	Ash (%)	0941	12.440	0.10000	11.950	0.39325	0.10936	125	1.25	005.05	0
005	Ash (%)	0955	12.450	0.10000	11.950	0.39325	0.10936	125	1.27	005.05	0
005	Ash (%)	0062	12.506	0.16400	11.950	0.39325	0.10936	125	1.41	005.00	0
005	Ash (%)	0100	12.530	0.14000	11.950	0.39325	0.10936	125	1.48	005.00	0
005	Ash (%)	0242	12.540	0.00000	11.950	0.39325	0.10936	125	1.50	005.00	0
005	Ash (%)	0685	12.575	0.05000	11.950	0.39325	0.10936	125	1.59	005.05	0
005	Ash (%)	0910	12.600	0.20000	11.950	0.39325	0.10936	125	1.65	005.99	0
005	Ash (%)	2129	12.695	0.13000	11.950	0.39325	0.10936	125	1.90	005.05	0
005	Ash (%)	0123	13.080	0.08000	11.950	0.39325	0.10936	125	2.87	005.00	0
005	Ash (%)	0297	13.113	0.54600	11.950	0.39325	0.10936	125	2.96	005.11	0
005	Ash (%)	2120	19.465	1.1900	11.950	0.39325	0.10936	125	19.11	005.11	2
005	Ash (%)	2121	19.480	0.60000	11.950	0.39325	0.10936	125	19.15	005.11	2
006	Total sugars (%)	2195	2.7700	0.06000	4.6263	1.6119	0.29250	4		006.00	0
006	Total sugars (%)	0910	3.8900	0.50000	4.6263	1.6119	0.29250	4		006.00	0
006	Total sugars (%)	0956	5.4500	0.50000	4.6263	1.6119	0.29250	4		006.99	0
006	Total sugars (%)	0407	6.3950	0.11000	4.6263	1.6119	0.29250	4		006.01	0
008	Fiber (%)	0910	11.545	1.2100	13.862	0.93436	0.28706	52	-2.48	008.08	0
008	Fiber (%)	0956	11.850	0.10000	13.862	0.93436	0.28706	52	-2.15	008.08	0
008	Fiber (%)	0590	12.300	0.00000	13.862	0.93436	0.28706	52	-1.67	008.02	0
008	Fiber (%)	0581	12.455	0.19000	13.862	0.93436	0.28706	52	-1.51	008.08	0
008	Fiber (%)	0035	12.625	0.09000	13.862	0.93436	0.28706	52	-1.32	008.08	0
008	Fiber (%)	0918	12.634	0.09500	13.862	0.93436	0.28706	52	-1.31	008.08	0
008	Fiber (%)	0354	12.790	0.10000	13.862	0.93436	0.28706	52	-1.15	008.08	0
008	Fiber (%)	0294	12.850	0.30000	13.862	0.93436	0.28706	52	-1.08	008.08	0
008	Fiber (%)	0675	12.920	0.00000	13.862	0.93436	0.28706	52	-1.01	008.02	0
008	Fiber (%)	0878	12.920	0.10000	13.862	0.93436	0.28706	52	-1.01	008.08	0
008	Fiber (%)	0049	13.190	0.36000	13.862	0.93436	0.28706	52	-0.72	008.08	0
008	Fiber (%)	0208	13.200	0.40000	13.862	0.93436	0.28706	52	-0.71	008.02	0
008	Fiber (%)	0889	13.220	0.04000	13.862	0.93436	0.28706	52	-0.69	008.08	0
008	Fiber (%)	0504	13.270	0.66000	13.862	0.93436	0.28706	52	-0.63	008.02	0
008	Fiber (%)	2188	13.395	0.29000	13.862	0.93436	0.28706	52	-0.50	008.08	0
008	Fiber (%)	0939	13.450	0.30000	13.862	0.93436	0.28706	52	-0.44	008.02	0
008	Fiber (%)	0278	13.450	0.10000	13.862	0.93436	0.28706	52	-0.44	008.08	0
008	Fiber (%)	0358	13.460	0.24000	13.862	0.93436	0.28706	52	-0.43	008.08	0
008	Fiber (%)	0263	13.498	0.11660	13.862	0.93436	0.28706	52	-0.39	008.08	0
008	Fiber (%)	2196	13.500	0.00000	13.862	0.93436	0.28706	52	-0.39	008.08	0
008	Fiber (%)	0357	13.600	0.00000	13.862	0.93436	0.28706	52	-0.28	008.08	0
008	Fiber (%)	0004	13.625	0.05000	13.862	0.93436	0.28706	52	-0.25	008.08	0
008	Fiber (%)	0413	13.650	0.90000	13.862	0.93436	0.28706	52	-0.23	008.08	0
008	Fiber (%)	0968	13.725	0.35400	13.862	0.93436	0.28706	52	-0.15	008.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
008	Fiber (%)	2146	13.730	0.86000	13.862	0.93436	0.28706	52	-0.14	008.08	0
008	Fiber (%)	0297	13.850	0.06000	13.862	0.93436	0.28706	52	-0.01	008.08	0
008	Fiber (%)	0723	13.944	0.00000	13.862	0.93436	0.28706	52	0.09	008.08	0
008	Fiber (%)	0407	14.074	0.21040	13.862	0.93436	0.28706	52	0.23	008.08	0
008	Fiber (%)	0948	14.110	0.26000	13.862	0.93436	0.28706	52	0.27	008.02	0
008	Fiber (%)	0098	14.120	0.14000	13.862	0.93436	0.28706	52	0.28	008.02	0
008	Fiber (%)	0001	14.135	0.09000	13.862	0.93436	0.28706	52	0.29	008.08	0
008	Fiber (%)	2129	14.170	0.86000	13.862	0.93436	0.28706	52	0.33	008.08	0
008	Fiber (%)	0353	14.260	0.06000	13.862	0.93436	0.28706	52	0.43	008.02	0
008	Fiber (%)	0510	14.300	0.40000	13.862	0.93436	0.28706	52	0.47	008.08	0
008	Fiber (%)	2192	14.320	1.4800	13.862	0.93436	0.28706	52	0.49	008.02	0
008	Fiber (%)	0190	14.340	0.40000	13.862	0.93436	0.28706	52	0.51	008.08	0
008	Fiber (%)	0646	14.355	0.41000	13.862	0.93436	0.28706	52	0.53	008.08	0
008	Fiber (%)	0226	14.385	0.27000	13.862	0.93436	0.28706	52	0.56	008.02	0
008	Fiber (%)	0693	14.416	0.23800	13.862	0.93436	0.28706	52	0.59	008.08	0
008	Fiber (%)	0848	14.425	0.01000	13.862	0.93436	0.28706	52	0.60	008.08	0
008	Fiber (%)	0870	14.461	0.45220	13.862	0.93436	0.28706	52	0.64	008.08	0
008	Fiber (%)	0689	14.600	0.40000	13.862	0.93436	0.28706	52	0.79	008.02	0
008	Fiber (%)	0037	14.600	0.40000	13.862	0.93436	0.28706	52	0.79	008.08	0
008	Fiber (%)	0309	14.606	0.01090	13.862	0.93436	0.28706	52	0.80	008.02	0
008	Fiber (%)	0187	14.655	0.03000	13.862	0.93436	0.28706	52	0.85	008.02	0
008	Fiber (%)	0035	14.805	0.45000	13.862	0.93436	0.28706	52	1.01	008.99	0
008	Fiber (%)	0726	15.135	0.11000	13.862	0.93436	0.28706	52	1.36	008.08	0
008	Fiber (%)	0345	15.155	0.09000	13.862	0.93436	0.28706	52	1.38	008.02	0
008	Fiber (%)	2195	15.350	0.06000	13.862	0.93436	0.28706	52	1.59	008.08	0
008	Fiber (%)	0083	15.540	0.12000	13.862	0.93436	0.28706	52	1.80	008.08	0
008	Fiber (%)	0045	15.650	0.50000	13.862	0.93436	0.28706	52	1.91	008.08	0
008	Fiber (%)	0536	16.140	0.56000	13.862	0.93436	0.28706	52	2.44	008.08	0
008	Fiber (%)	0609	35.400	7.5200	13.862	0.93436	0.28706	52	23.05	008.02	2
009	Fiber (%)	0910	21.000	0.00000	23.709	1.8789	0.42317	49	-1.44	009.07	0
009	Fiber (%)	0956	21.100	0.20000	23.709	1.8789	0.42317	49	-1.39	009.09	0
009	Fiber (%)	0278	21.350	0.90000	23.709	1.8789	0.42317	49	-1.26	009.09	0
009	Fiber (%)	0049	21.505	0.85000	23.709	1.8789	0.42317	49	-1.17	009.09	0
009	Fiber (%)	2196	21.800	0.00000	23.709	1.8789	0.42317	49	-1.02	009.09	0
009	Fiber (%)	0939	21.975	0.13000	23.709	1.8789	0.42317	49	-0.92	009.09	0
009	Fiber (%)	0187	22.100	0.08000	23.709	1.8789	0.42317	49	-0.86	009.07	0
009	Fiber (%)	0357	22.200	1.0000	23.709	1.8789	0.42317	49	-0.80	009.09	0
009	Fiber (%)	0263	22.346	0.08250	23.709	1.8789	0.42317	49	-0.73	009.09	0
009	Fiber (%)	0889	22.370	0.04000	23.709	1.8789	0.42317	49	-0.71	009.09	0
009	Fiber (%)	0870	22.464	1.4989	23.709	1.8789	0.42317	49	-0.66	009.09	0
009	Fiber (%)	0294	22.500	0.00000	23.709	1.8789	0.42317	49	-0.64	009.09	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
009	Fiber (%)	0918	22.508	1.2489	23.709	1.8789	0.42317	49	-0.64	009.09	0
009	Fiber (%)	0358	22.560	1.2400	23.709	1.8789	0.42317	49	-0.61	009.09	0
009	Fiber (%)	0878	22.635	0.17000	23.709	1.8789	0.42317	49	-0.57	009.09	0
009	Fiber (%)	0098	22.650	0.90000	23.709	1.8789	0.42317	49	-0.56	009.07	0
009	Fiber (%)	0675	22.705	0.03000	23.709	1.8789	0.42317	49	-0.53	009.07	0
009	Fiber (%)	0693	22.711	0.13000	23.709	1.8789	0.42317	49	-0.53	009.09	0
009	Fiber (%)	0581	22.715	0.19000	23.709	1.8789	0.42317	49	-0.53	009.09	0
009	Fiber (%)	0001	22.770	0.20000	23.709	1.8789	0.42317	49	-0.50	009.09	0
009	Fiber (%)	0037	23.085	0.53000	23.709	1.8789	0.42317	49	-0.33	009.09	0
009	Fiber (%)	0413	23.200	0.40000	23.709	1.8789	0.42317	49	-0.27	009.09	0
009	Fiber (%)	0646	23.200	0.32000	23.709	1.8789	0.42317	49	-0.27	009.09	0
009	Fiber (%)	0353	23.285	0.39000	23.709	1.8789	0.42317	49	-0.23	009.07	0
009	Fiber (%)	2195	23.350	0.04000	23.709	1.8789	0.42317	49	-0.19	009.09	0
009	Fiber (%)	0045	23.500	0.20000	23.709	1.8789	0.42317	49	-0.11	009.09	0
009	Fiber (%)	0354	23.505	0.19000	23.709	1.8789	0.42317	49	-0.11	009.09	0
009	Fiber (%)	2129	23.510	0.70000	23.709	1.8789	0.42317	49	-0.11	009.09	0
009	Fiber (%)	0590	23.650	1.1000	23.709	1.8789	0.42317	49	-0.03	009.07	0
009	Fiber (%)	0297	23.690	0.14000	23.709	1.8789	0.42317	49	-0.01	009.09	0
009	Fiber (%)	0638	23.700	0.00000	23.709	1.8789	0.42317	49	0.00	009.09	0
009	Fiber (%)	0723	23.737	0.00000	23.709	1.8789	0.42317	49	0.01	009.09	0
009	Fiber (%)	0948	23.930	0.28000	23.709	1.8789	0.42317	49	0.12	009.09	0
009	Fiber (%)	0226	23.975	0.13000	23.709	1.8789	0.42317	49	0.14	009.07	0
009	Fiber (%)	0689	24.500	0.20000	23.709	1.8789	0.42317	49	0.42	009.07	0
009	Fiber (%)	2192	24.740	0.26000	23.709	1.8789	0.42317	49	0.55	009.07	0
009	Fiber (%)	0407	25.224	0.48580	23.709	1.8789	0.42317	49	0.81	009.09	0
009	Fiber (%)	2188	25.240	0.40000	23.709	1.8789	0.42317	49	0.82	009.09	0
009	Fiber (%)	0083	25.450	1.0400	23.709	1.8789	0.42317	49	0.93	009.09	0
009	Fiber (%)	0968	25.810	0.40100	23.709	1.8789	0.42317	49	1.12	009.09	0
009	Fiber (%)	0190	26.025	0.13000	23.709	1.8789	0.42317	49	1.23	009.09	0
009	Fiber (%)	0510	26.050	0.70000	23.709	1.8789	0.42317	49	1.25	009.09	0
009	Fiber (%)	0309	26.267	0.00800	23.709	1.8789	0.42317	49	1.36	009.07	0
009	Fiber (%)	2146	26.500	1.2400	23.709	1.8789	0.42317	49	1.49	009.09	0
009	Fiber (%)	0208	26.650	0.10000	23.709	1.8789	0.42317	49	1.57	009.07	0
009	Fiber (%)	0726	26.660	0.08000	23.709	1.8789	0.42317	49	1.57	009.09	0
009	Fiber (%)	0848	26.660	0.08000	23.709	1.8789	0.42317	49	1.57	009.09	0
009	Fiber (%)	0536	26.760	1.8000	23.709	1.8789	0.42317	49	1.62	009.09	0
009	Fiber (%)	0345	27.150	0.50000	23.709	1.8789	0.42317	49	1.83	009.07	0
009	Fiber (%)	0504	23.170	2.3400	23.709	1.8789	0.42317	49	-0.29	009.04	1
009	Fiber (%)	0609	50.450	0.16000	23.709	1.8789	0.42317	49	14.23	009.07	2
010	Moisture (%)	0948	7.0900	0.04000	7.6266	0.33194	0.08174	23	-1.62	010.99	0
010	Moisture (%)	0208	7.3050	0.09000	7.6266	0.33194	0.08174	23	-0.97	010.03	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
010	Moisture (%)	2120	7.3150	0.01000	7.6266	0.33194	0.08174	23	-0.94	010.11	0
010	Moisture (%)	2089	7.3200	0.44000	7.6266	0.33194	0.08174	23	-0.92	010.99	0
010	Moisture (%)	0939	7.3950	0.41000	7.6266	0.33194	0.08174	23	-0.70	010.99	0
010	Moisture (%)	0692	7.4000	0.00000	7.6266	0.33194	0.08174	23	-0.68	010.99	0
010	Moisture (%)	0884	7.4000	0.00000	7.6266	0.33194	0.08174	23	-0.68	010.99	0
010	Moisture (%)	2129	7.4450	0.07000	7.6266	0.33194	0.08174	23	-0.55	010.99	0
010	Moisture (%)	2121	7.5150	0.03000	7.6266	0.33194	0.08174	23	-0.34	010.11	0
010	Moisture (%)	2043	7.5300	0.06000	7.6266	0.33194	0.08174	23	-0.29	010.99	0
010	Moisture (%)	0921	7.5400	0.00000	7.6266	0.33194	0.08174	23	-0.26	010.99	0
010	Moisture (%)	0889	7.5850	0.09000	7.6266	0.33194	0.08174	23	-0.13	010.11	0
010	Moisture (%)	0652	7.6000	0.00000	7.6266	0.33194	0.08174	23	-0.08	010.99	0
010	Moisture (%)	0190	7.6350	0.05000	7.6266	0.33194	0.08174	23	0.03	010.99	0
010	Moisture (%)	2178	7.6700	0.02000	7.6266	0.33194	0.08174	23	0.13	010.99	0
010	Moisture (%)	2076	7.7000	0.10000	7.6266	0.33194	0.08174	23	0.22	010.99	0
010	Moisture (%)	2105	7.7450	0.05000	7.6266	0.33194	0.08174	23	0.36	010.11	0
010	Moisture (%)	2150	7.7550	0.15000	7.6266	0.33194	0.08174	23	0.39	010.99	0
010	Moisture (%)	0726	7.9300	0.02000	7.6266	0.33194	0.08174	23	0.91	010.99	0
010	Moisture (%)	0152	8.2000	0.20000	7.6266	0.33194	0.08174	23	1.73	010.11	0
010	Moisture (%)	0952	8.2150	0.01000	7.6266	0.33194	0.08174	23	1.77	010.99	0
010	Moisture (%)	2127	9.0000	0.00000	7.6266	0.33194	0.08174	23	4.14	010.99	0
010	Moisture (%)	0405	9.0800	0.04000	7.6266	0.33194	0.08174	23	4.38	010.99	0
010	Moisture (%)	0028	6.2150	1.5500	7.6266	0.33194	0.08174	23	-4.25	010.99	1
011	Loss on Drying (%)	0536	6.1650	0.17000	8.2114	0.29858	0.06575	56	-6.85	011.01	0
011	Loss on Drying (%)	0345	6.8950	0.07000	8.2114	0.29858	0.06575	56	-4.41	011.01	0
011	Loss on Drying (%)	0874	7.3750	0.07000	8.2114	0.29858	0.06575	56	-2.80	011.01	0
011	Loss on Drying (%)	0574	7.4950	0.05000	8.2114	0.29858	0.06575	56	-2.40	011.01	0
011	Loss on Drying (%)	0175	7.6500	0.10000	8.2114	0.29858	0.06575	56	-1.88	011.01	0
011	Loss on Drying (%)	0859	7.6505	0.03300	8.2114	0.29858	0.06575	56	-1.88	011.01	0
011	Loss on Drying (%)	0108	7.7950	0.31000	8.2114	0.29858	0.06575	56	-1.39	011.01	0
011	Loss on Drying (%)	0598	7.8100	0.04000	8.2114	0.29858	0.06575	56	-1.34	011.01	0
011	Loss on Drying (%)	2188	7.8450	0.01000	8.2114	0.29858	0.06575	56	-1.23	011.01	0
011	Loss on Drying (%)	0723	7.8700	0.00000	8.2114	0.29858	0.06575	56	-1.14	011.01	0
011	Loss on Drying (%)	0062	7.9395	0.03700	8.2114	0.29858	0.06575	56	-0.91	011.01	0
011	Loss on Drying (%)	0953	7.9650	0.25000	8.2114	0.29858	0.06575	56	-0.83	011.01	0
011	Loss on Drying (%)	0226	7.9700	0.08000	8.2114	0.29858	0.06575	56	-0.81	011.01	0
011	Loss on Drying (%)	0968	8.0180	0.01000	8.2114	0.29858	0.06575	56	-0.65	011.01	0
011	Loss on Drying (%)	2181	8.0300	0.02000	8.2114	0.29858	0.06575	56	-0.61	011.01	0
011	Loss on Drying (%)	0208	8.0550	0.01000	8.2114	0.29858	0.06575	56	-0.52	011.01	0
011	Loss on Drying (%)	0942	8.1200	0.36000	8.2114	0.29858	0.06575	56	-0.31	011.02	0
011	Loss on Drying (%)	0541	8.1350	0.15000	8.2114	0.29858	0.06575	56	-0.26	011.01	0
011	Loss on Drying (%)	2073	8.1400	0.02000	8.2114	0.29858	0.06575	56	-0.24	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
011	Loss on Drying (%)	0539	8.1500	0.16000	8.2114	0.29858	0.06575	56	-0.21	011.01	0
011	Loss on Drying (%)	0100	8.1700	0.02000	8.2114	0.29858	0.06575	56	-0.14	011.01	0
011	Loss on Drying (%)	0653	8.1700	0.02000	8.2114	0.29858	0.06575	56	-0.14	011.01	0
011	Loss on Drying (%)	0881	8.1825	0.03700	8.2114	0.29858	0.06575	56	-0.10	011.01	0
011	Loss on Drying (%)	0520	8.2000	0.00000	8.2114	0.29858	0.06575	56	-0.04	011.01	0
011	Loss on Drying (%)	2196	8.2000	0.00000	8.2114	0.29858	0.06575	56	-0.04	011.01	0
011	Loss on Drying (%)	2006	8.2200	0.04000	8.2114	0.29858	0.06575	56	0.03	011.01	0
011	Loss on Drying (%)	0646	8.2250	0.13000	8.2114	0.29858	0.06575	56	0.05	011.01	0
011	Loss on Drying (%)	2193	8.2300	0.14000	8.2114	0.29858	0.06575	56	0.06	011.99	0
011	Loss on Drying (%)	0309	8.2350	0.09000	8.2114	0.29858	0.06575	56	0.08	011.01	0
011	Loss on Drying (%)	0425	8.2550	0.05000	8.2114	0.29858	0.06575	56	0.15	011.02	0
011	Loss on Drying (%)	0682	8.2800	0.00000	8.2114	0.29858	0.06575	56	0.23	011.01	0
011	Loss on Drying (%)	0242	8.2900	0.08000	8.2114	0.29858	0.06575	56	0.26	011.01	0
011	Loss on Drying (%)	0510	8.3000	0.00000	8.2114	0.29858	0.06575	56	0.30	011.01	0
011	Loss on Drying (%)	0098	8.3050	0.03000	8.2114	0.29858	0.06575	56	0.31	011.01	0
011	Loss on Drying (%)	0955	8.3150	0.07000	8.2114	0.29858	0.06575	56	0.35	011.01	0
011	Loss on Drying (%)	2168	8.3150	0.07000	8.2114	0.29858	0.06575	56	0.35	011.01	0
011	Loss on Drying (%)	0407	8.3373	0.03200	8.2114	0.29858	0.06575	56	0.42	011.01	0
011	Loss on Drying (%)	0202	8.3400	0.02000	8.2114	0.29858	0.06575	56	0.43	011.01	0
011	Loss on Drying (%)	0354	8.3450	0.01000	8.2114	0.29858	0.06575	56	0.45	011.01	0
011	Loss on Drying (%)	0144	8.3500	0.12000	8.2114	0.29858	0.06575	56	0.46	011.01	0
011	Loss on Drying (%)	0870	8.3542	0.04930	8.2114	0.29858	0.06575	56	0.48	011.01	0
011	Loss on Drying (%)	0675	8.3750	0.15000	8.2114	0.29858	0.06575	56	0.55	011.01	0
011	Loss on Drying (%)	0589	8.3800	0.04000	8.2114	0.29858	0.06575	56	0.56	011.01	0
011	Loss on Drying (%)	0229	8.3860	0.01200	8.2114	0.29858	0.06575	56	0.58	011.01	0
011	Loss on Drying (%)	0294	8.4500	0.10000	8.2114	0.29858	0.06575	56	0.80	011.01	0
011	Loss on Drying (%)	0123	8.4600	0.00000	8.2114	0.29858	0.06575	56	0.83	011.01	0
011	Loss on Drying (%)	0171	8.4600	0.00000	8.2114	0.29858	0.06575	56	0.83	011.01	0
011	Loss on Drying (%)	0685	8.5000	0.00000	8.2114	0.29858	0.06575	56	0.97	011.01	0
011	Loss on Drying (%)	0233	8.5400	0.06000	8.2114	0.29858	0.06575	56	1.10	011.01	0
011	Loss on Drying (%)	0876	8.5450	0.09000	8.2114	0.29858	0.06575	56	1.12	011.01	0
011	Loss on Drying (%)	2022	8.5450	0.01000	8.2114	0.29858	0.06575	56	1.12	011.01	0
011	Loss on Drying (%)	0848	8.5850	0.01000	8.2114	0.29858	0.06575	56	1.25	011.01	0
011	Loss on Drying (%)	0205	8.5930	0.01800	8.2114	0.29858	0.06575	56	1.28	011.01	0
011	Loss on Drying (%)	0263	8.6648	0.00360	8.2114	0.29858	0.06575	56	1.52	011.01	0
011	Loss on Drying (%)	0358	8.8150	0.19000	8.2114	0.29858	0.06575	56	2.02	011.01	0
011	Loss on Drying (%)	0004	8.8200	0.04000	8.2114	0.29858	0.06575	56	2.04	011.01	0
011	Loss on Drying (%)	0857	7.9500	0.50000	8.2114	0.29858	0.06575	56	-0.88	011.99	1
011	Loss on Drying (%)	0511	8.0350	0.45000	8.2114	0.29858	0.06575	56	-0.59	011.01	1
012	Starch (%)	0918	11.409	0.84500	16.832	1.4192	0.29657	36	-3.82	012.01	0
012	Starch (%)	2066	13.985	1.0700	16.832	1.4192	0.29657	36	-2.01	012.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
012	Starch (%)	0885	14.636	0.11160	16.832	1.4192	0.29657	36	-1.55	012.03	0
012	Starch (%)	0160	15.215	0.33000	16.832	1.4192	0.29657	36	-1.14	012.04	0
012	Starch (%)	0870	15.489	0.92270	16.832	1.4192	0.29657	36	-0.95	012.01	0
012	Starch (%)	0045	15.500	0.40000	16.832	1.4192	0.29657	36	-0.94	012.01	0
012	Starch (%)	0297	15.520	0.38000	16.832	1.4192	0.29657	36	-0.92	012.11	0
012	Starch (%)	0098	15.600	0.20000	16.832	1.4192	0.29657	36	-0.87	012.04	0
012	Starch (%)	2195	15.630	0.44000	16.832	1.4192	0.29657	36	-0.85	012.00	0
012	Starch (%)	0956	16.000	0.40000	16.832	1.4192	0.29657	36	-0.59	012.01	0
012	Starch (%)	0407	16.020	0.14000	16.832	1.4192	0.29657	36	-0.57	012.03	0
012	Starch (%)	0723	16.060	0.00000	16.832	1.4192	0.29657	36	-0.54	012.00	0
012	Starch (%)	0353	16.075	0.31000	16.832	1.4192	0.29657	36	-0.53	012.04	0
012	Starch (%)	0278	16.200	0.20000	16.832	1.4192	0.29657	36	-0.45	012.04	0
012	Starch (%)	0689	16.300	0.40000	16.832	1.4192	0.29657	36	-0.37	012.00	0
012	Starch (%)	0889	16.375	0.41000	16.832	1.4192	0.29657	36	-0.32	012.11	0
012	Starch (%)	0693	16.652	1.0740	16.832	1.4192	0.29657	36	-0.13	012.01	0
012	Starch (%)	0510	16.800	0.40000	16.832	1.4192	0.29657	36	-0.02	012.04	0
012	Starch (%)	0297	16.805	0.01000	16.832	1.4192	0.29657	36	-0.02	012.03	0
012	Starch (%)	0208	17.100	0.40000	16.832	1.4192	0.29657	36	0.19	012.04	0
012	Starch (%)	2196	17.150	0.00000	16.832	1.4192	0.29657	36	0.22	012.00	0
012	Starch (%)	0610	17.500	0.00000	16.832	1.4192	0.29657	36	0.47	012.00	0
012	Starch (%)	0939	17.635	0.05000	16.832	1.4192	0.29657	36	0.57	012.00	0
012	Starch (%)	0683	17.695	0.27000	16.832	1.4192	0.29657	36	0.61	012.00	0
012	Starch (%)	2006	17.695	0.39000	16.832	1.4192	0.29657	36	0.61	012.00	0
012	Starch (%)	0354	17.800	0.02000	16.832	1.4192	0.29657	36	0.68	012.00	0
012	Starch (%)	0638	17.800	0.00000	16.832	1.4192	0.29657	36	0.68	012.00	0
012	Starch (%)	2188	17.850	0.26000	16.832	1.4192	0.29657	36	0.72	012.00	0
012	Starch (%)	2181	17.930	0.00000	16.832	1.4192	0.29657	36	0.77	012.00	0
012	Starch (%)	0609	18.060	0.00000	16.832	1.4192	0.29657	36	0.87	012.00	0
012	Starch (%)	0948	18.065	0.13000	16.832	1.4192	0.29657	36	0.87	012.04	0
012	Starch (%)	2009	18.416	0.02230	16.832	1.4192	0.29657	36	1.12	012.00	0
012	Starch (%)	0910	18.500	1.0000	16.832	1.4192	0.29657	36	1.18	012.00	0
012	Starch (%)	0878	18.600	0.00000	16.832	1.4192	0.29657	36	1.25	012.00	0
012	Starch (%)	0889	18.846	0.09100	16.832	1.4192	0.29657	36	1.42	012.03	0
012	Starch (%)	2022	28.460	0.00000	16.832	1.4192	0.29657	36	8.19	012.99	0
013	Fat (%)	0353	0.90500	0.25000	5.2230	1.3390	0.23654	45	-3.22	013.10	0
013	Fat (%)	0202	2.7450	0.05000	5.2230	1.3390	0.23654	45	-1.85	013.00	0
013	Fat (%)	0885	3.2272	0.09230	5.2230	1.3390	0.23654	45	-1.49	013.00	0
013	Fat (%)	2146	3.3750	0.13000	5.2230	1.3390	0.23654	45	-1.38	013.00	0
013	Fat (%)	2145	3.3900	0.00000	5.2230	1.3390	0.23654	45	-1.37	013.00	0
013	Fat (%)	0689	3.4500	0.10000	5.2230	1.3390	0.23654	45	-1.32	013.00	0
013	Fat (%)	0618	3.7518	0.17310	5.2230	1.3390	0.23654	45	-1.10	013.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
013	Fat (%)	0968	3.8495	0.00700	5.2230	1.3390	0.23654	45	-1.03	013.13	0
013	Fat (%)	0242	3.8850	0.51000	5.2230	1.3390	0.23654	45	-1.00	013.00	0
013	Fat (%)	0874	4.0300	0.04000	5.2230	1.3390	0.23654	45	-0.89	013.02	0
013	Fat (%)	0011	4.1350	0.31000	5.2230	1.3390	0.23654	45	-0.81	013.02	0
013	Fat (%)	0036	4.2167	0.50570	5.2230	1.3390	0.23654	45	-0.75	013.10	0
013	Fat (%)	0504	4.3900	0.34000	5.2230	1.3390	0.23654	45	-0.62	013.00	0
013	Fat (%)	2006	4.4550	0.21000	5.2230	1.3390	0.23654	45	-0.57	013.10	0
013	Fat (%)	0610	4.5050	0.11000	5.2230	1.3390	0.23654	45	-0.54	013.10	0
013	Fat (%)	0407	4.7031	0.21540	5.2230	1.3390	0.23654	45	-0.39	013.13	0
013	Fat (%)	0955	4.9600	0.00000	5.2230	1.3390	0.23654	45	-0.20	013.13	0
013	Fat (%)	0884	5.0500	0.10000	5.2230	1.3390	0.23654	45	-0.13	013.00	0
013	Fat (%)	0652	5.1000	0.40000	5.2230	1.3390	0.23654	45	-0.09	013.00	0
013	Fat (%)	0948	5.1350	0.09000	5.2230	1.3390	0.23654	45	-0.07	013.00	0
013	Fat (%)	0098	5.3150	0.05000	5.2230	1.3390	0.23654	45	0.07	013.02	0
013	Fat (%)	0910	5.3800	0.52000	5.2230	1.3390	0.23654	45	0.12	013.00	0
013	Fat (%)	0541	5.3900	0.36000	5.2230	1.3390	0.23654	45	0.12	013.00	0
013	Fat (%)	0027	5.4300	0.82000	5.2230	1.3390	0.23654	45	0.15	013.13	0
013	Fat (%)	0229	5.5950	0.41000	5.2230	1.3390	0.23654	45	0.28	013.00	0
013	Fat (%)	0618	5.6686	0.50590	5.2230	1.3390	0.23654	45	0.33	013.00	0
013	Fat (%)	0643	5.7200	0.46000	5.2230	1.3390	0.23654	45	0.37	013.02	0
013	Fat (%)	0921	5.7300	0.10000	5.2230	1.3390	0.23654	45	0.38	013.02	0
013	Fat (%)	0675	5.7600	0.16000	5.2230	1.3390	0.23654	45	0.40	013.02	0
013	Fat (%)	0309	5.8248	0.18100	5.2230	1.3390	0.23654	45	0.45	013.00	0
013	Fat (%)	0016	5.8900	0.10000	5.2230	1.3390	0.23654	45	0.50	013.02	0
013	Fat (%)	2073	5.9700	0.20000	5.2230	1.3390	0.23654	45	0.56	013.00	0
013	Fat (%)	0208	6.0550	0.11000	5.2230	1.3390	0.23654	45	0.62	013.02	0
013	Fat (%)	2150	6.1050	0.45000	5.2230	1.3390	0.23654	45	0.66	013.02	0
013	Fat (%)	0139	6.1550	0.29000	5.2230	1.3390	0.23654	45	0.70	013.00	0
013	Fat (%)	0638	6.3500	0.30000	5.2230	1.3390	0.23654	45	0.84	013.10	0
013	Fat (%)	0010	6.4000	0.40000	5.2230	1.3390	0.23654	45	0.88	013.02	0
013	Fat (%)	0870	6.4348	0.09750	5.2230	1.3390	0.23654	45	0.90	013.02	0
013	Fat (%)	2076	6.5282	0.36660	5.2230	1.3390	0.23654	45	0.97	013.00	0
013	Fat (%)	0682	6.5600	0.00000	5.2230	1.3390	0.23654	45	1.00	013.02	0
013	Fat (%)	0100	6.6100	0.30000	5.2230	1.3390	0.23654	45	1.04	013.02	0
013	Fat (%)	0676	6.7100	0.12000	5.2230	1.3390	0.23654	45	1.11	013.02	0
013	Fat (%)	0354	7.0200	0.14000	5.2230	1.3390	0.23654	45	1.34	013.02	0
013	Fat (%)	0171	7.1650	0.15000	5.2230	1.3390	0.23654	45	1.45	013.02	0
013	Fat (%)	0581	7.4900	0.42000	5.2230	1.3390	0.23654	45	1.69	013.13	0
015	Aluminum (mg / kg (ppm))	0186	207.00	12.000	315.69	58.426	17.052	11	-1.86	015.52	0
015	Aluminum (mg / kg (ppm))	0918	251.26	27.778	315.69	58.426	17.052	11	-1.10	015.43	0
015	Aluminum (mg / kg (ppm))	0169	256.00	2.0000	315.69	58.426	17.052	11	-1.02	015.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
015	Aluminum (mg / kg (ppm))	0510	297.00	2.0000	315.69	58.426	17.052	11	-0.32	015.43	0
015	Aluminum (mg / kg (ppm))	0049	314.25	9.8900	315.69	58.426	17.052	11	-0.02	015.41	0
015	Aluminum (mg / kg (ppm))	0297	330.00	70.000	315.69	58.426	17.052	11	0.24	015.43	0
015	Aluminum (mg / kg (ppm))	0560	335.55	12.900	315.69	58.426	17.052	11	0.34	015.42	0
015	Aluminum (mg / kg (ppm))	0353	353.00	4.0000	315.69	58.426	17.052	11	0.64	015.43	0
015	Aluminum (mg / kg (ppm))	0520	363.00	36.000	315.69	58.426	17.052	11	0.81	015.41	0
015	Aluminum (mg / kg (ppm))	0345	363.00	6.0000	315.69	58.426	17.052	11	0.81	015.43	0
015	Aluminum (mg / kg (ppm))	0171	381.50	5.0000	315.69	58.426	17.052	11	1.13	015.41	0
017	Boron (mg / kg (ppm))	0297	15.000	0.00000	19.005	1.5451	0.59608	12	-2.59	017.43	0
017	Boron (mg / kg (ppm))	0693	17.498	0.03300	19.005	1.5451	0.59608	12	-0.98	017.42	0
017	Boron (mg / kg (ppm))	0510	18.000	0.00000	19.005	1.5451	0.59608	12	-0.65	017.43	0
017	Boron (mg / kg (ppm))	0353	18.500	1.0000	19.005	1.5451	0.59608	12	-0.33	017.43	0
017	Boron (mg / kg (ppm))	0045	18.650	2.3000	19.005	1.5451	0.59608	12	-0.23	017.42	0
017	Boron (mg / kg (ppm))	0345	18.650	0.30000	19.005	1.5451	0.59608	12	-0.23	017.43	0
017	Boron (mg / kg (ppm))	0560	18.785	0.37000	19.005	1.5451	0.59608	12	-0.14	017.42	0
017	Boron (mg / kg (ppm))	0918	19.060	0.40000	19.005	1.5451	0.59608	12	0.04	017.43	0
017	Boron (mg / kg (ppm))	0049	19.705	0.05000	19.005	1.5451	0.59608	12	0.45	017.41	0
017	Boron (mg / kg (ppm))	0358	20.235	0.13000	19.005	1.5451	0.59608	12	0.80	017.41	0
017	Boron (mg / kg (ppm))	0229	20.965	2.4700	19.005	1.5451	0.59608	12	1.27	017.41	0
017	Boron (mg / kg (ppm))	0294	21.650	0.10000	19.005	1.5451	0.59608	12	1.71	017.42	0
017	Boron (mg / kg (ppm))	2053	23.310	4.6000	19.005	1.5451	0.59608	12	2.79	017.42	1
019	Calcium (%)	0590	1.5800	0.00000	2.1819	0.12488	0.04788	122	-4.82	019.99	0
019	Calcium (%)	2188	1.7100	0.00000	2.1819	0.12488	0.04788	122	-3.78	019.08	0
019	Calcium (%)	2066	1.7800	0.00000	2.1819	0.12488	0.04788	122	-3.22	019.53	0
019	Calcium (%)	0889	1.8650	0.02030	2.1819	0.12488	0.04788	122	-2.54	019.99	0
019	Calcium (%)	2192	1.9400	0.02000	2.1819	0.12488	0.04788	122	-1.94	019.43	0
019	Calcium (%)	0536	1.9500	0.22000	2.1819	0.12488	0.04788	122	-1.86	019.31	0
019	Calcium (%)	0043	1.9650	0.03000	2.1819	0.12488	0.04788	122	-1.74	019.43	0
019	Calcium (%)	0144	1.9800	0.16000	2.1819	0.12488	0.04788	122	-1.62	019.41	0
019	Calcium (%)	0510	1.9950	0.05000	2.1819	0.12488	0.04788	122	-1.50	019.43	0
019	Calcium (%)	0042	2.0050	0.11000	2.1819	0.12488	0.04788	122	-1.42	019.42	0
019	Calcium (%)	0692	2.0050	0.07000	2.1819	0.12488	0.04788	122	-1.42	019.42	0
019	Calcium (%)	0154	2.0059	0.11560	2.1819	0.12488	0.04788	122	-1.41	019.52	0
019	Calcium (%)	0511	2.0200	0.24000	2.1819	0.12488	0.04788	122	-1.30	019.41	0
019	Calcium (%)	2181	2.0250	0.05000	2.1819	0.12488	0.04788	122	-1.26	019.00	0
019	Calcium (%)	0233	2.0250	0.13000	2.1819	0.12488	0.04788	122	-1.26	019.31	0
019	Calcium (%)	0036	2.0379	0.05500	2.1819	0.12488	0.04788	122	-1.15	019.43	0
019	Calcium (%)	0693	2.0430	0.03600	2.1819	0.12488	0.04788	122	-1.11	019.42	0
019	Calcium (%)	2188	2.0550	0.07000	2.1819	0.12488	0.04788	122	-1.02	019.31	0
019	Calcium (%)	0098	2.0550	0.01000	2.1819	0.12488	0.04788	122	-1.02	019.41	0
019	Calcium (%)	0045	2.0550	0.07000	2.1819	0.12488	0.04788	122	-1.02	019.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	0629	2.0650	0.01000	2.1819	0.12488	0.04788	122	-0.94	019.08	0
019	Calcium (%)	2022	2.0650	0.01000	2.1819	0.12488	0.04788	122	-0.94	019.31	0
019	Calcium (%)	0609	2.0650	0.07000	2.1819	0.12488	0.04788	122	-0.94	019.32	0
019	Calcium (%)	0019	2.0650	0.11000	2.1819	0.12488	0.04788	122	-0.94	019.41	0
019	Calcium (%)	0175	2.0700	0.02000	2.1819	0.12488	0.04788	122	-0.90	019.31	0
019	Calcium (%)	0848	2.0800	0.00000	2.1819	0.12488	0.04788	122	-0.82	019.41	0
019	Calcium (%)	0723	2.0923	0.00000	2.1819	0.12488	0.04788	122	-0.72	019.31	0
019	Calcium (%)	0205	2.0950	0.05000	2.1819	0.12488	0.04788	122	-0.70	019.42	0
019	Calcium (%)	0676	2.0950	0.05000	2.1819	0.12488	0.04788	122	-0.70	019.99	0
019	Calcium (%)	0560	2.1003	0.01450	2.1819	0.12488	0.04788	122	-0.65	019.42	0
019	Calcium (%)	0208	2.1040	0.16000	2.1819	0.12488	0.04788	122	-0.62	019.31	0
019	Calcium (%)	0675	2.1050	0.01000	2.1819	0.12488	0.04788	122	-0.62	019.43	0
019	Calcium (%)	0297	2.1100	0.10000	2.1819	0.12488	0.04788	122	-0.58	019.43	0
019	Calcium (%)	0653	2.1120	0.00200	2.1819	0.12488	0.04788	122	-0.56	019.31	0
019	Calcium (%)	0011	2.1185	0.08710	2.1819	0.12488	0.04788	122	-0.51	019.41	0
019	Calcium (%)	0910	2.1200	0.10000	2.1819	0.12488	0.04788	122	-0.50	019.41	0
019	Calcium (%)	2145	2.1200	0.00000	2.1819	0.12488	0.04788	122	-0.50	019.41	0
019	Calcium (%)	0171	2.1250	0.03000	2.1819	0.12488	0.04788	122	-0.46	019.41	0
019	Calcium (%)	0294	2.1250	0.03000	2.1819	0.12488	0.04788	122	-0.46	019.42	0
019	Calcium (%)	0948	2.1270	0.02800	2.1819	0.12488	0.04788	122	-0.44	019.31	0
019	Calcium (%)	0018	2.1350	0.01000	2.1819	0.12488	0.04788	122	-0.38	019.43	0
019	Calcium (%)	0345	2.1350	0.03000	2.1819	0.12488	0.04788	122	-0.38	019.43	0
019	Calcium (%)	0878	2.1400	0.04000	2.1819	0.12488	0.04788	122	-0.34	019.00	0
019	Calcium (%)	0870	2.1435	0.14300	2.1819	0.12488	0.04788	122	-0.31	019.42	0
019	Calcium (%)	2109	2.1470	0.00200	2.1819	0.12488	0.04788	122	-0.28	019.00	0
019	Calcium (%)	0878	2.1500	0.10000	2.1819	0.12488	0.04788	122	-0.26	019.41	0
019	Calcium (%)	0003	2.1550	0.01000	2.1819	0.12488	0.04788	122	-0.22	019.41	0
019	Calcium (%)	0208	2.1570	0.01200	2.1819	0.12488	0.04788	122	-0.20	019.41	0
019	Calcium (%)	0968	2.1595	0.00100	2.1819	0.12488	0.04788	122	-0.18	019.43	0
019	Calcium (%)	0682	2.1600	0.00000	2.1819	0.12488	0.04788	122	-0.18	019.43	0
019	Calcium (%)	0508	2.1619	0.00530	2.1819	0.12488	0.04788	122	-0.16	019.43	0
019	Calcium (%)	0229	2.1635	0.01700	2.1819	0.12488	0.04788	122	-0.15	019.41	0
019	Calcium (%)	0884	2.1650	0.01000	2.1819	0.12488	0.04788	122	-0.14	019.31	0
019	Calcium (%)	0139	2.1650	0.01000	2.1819	0.12488	0.04788	122	-0.14	019.41	0
019	Calcium (%)	0278	2.1650	0.05000	2.1819	0.12488	0.04788	122	-0.14	019.42	0
019	Calcium (%)	0263	2.1683	0.00220	2.1819	0.12488	0.04788	122	-0.11	019.42	0
019	Calcium (%)	0037	2.1700	0.08000	2.1819	0.12488	0.04788	122	-0.09	019.42	0
019	Calcium (%)	0242	2.1700	0.04000	2.1819	0.12488	0.04788	122	-0.09	019.99	0
019	Calcium (%)	0590	2.1750	0.05000	2.1819	0.12488	0.04788	122	-0.05	019.08	0
019	Calcium (%)	0939	2.1800	0.06000	2.1819	0.12488	0.04788	122	-0.01	019.31	0
019	Calcium (%)	0010	2.1800	0.06000	2.1819	0.12488	0.04788	122	-0.01	019.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	0407	2.1822	0.00800	2.1819	0.12488	0.04788	122	0.00	019.41	0
019	Calcium (%)	0955	2.1850	0.07000	2.1819	0.12488	0.04788	122	0.03	019.44	0
019	Calcium (%)	0202	2.1935	0.02300	2.1819	0.12488	0.04788	122	0.09	019.43	0
019	Calcium (%)	0354	2.1950	0.01000	2.1819	0.12488	0.04788	122	0.11	019.31	0
019	Calcium (%)	0098	2.1950	0.01000	2.1819	0.12488	0.04788	122	0.11	019.43	0
019	Calcium (%)	2009	2.1996	0.00310	2.1819	0.12488	0.04788	122	0.14	019.08	0
019	Calcium (%)	2196	2.2000	0.00000	2.1819	0.12488	0.04788	122	0.15	019.08	0
019	Calcium (%)	0047	2.2000	0.00000	2.1819	0.12488	0.04788	122	0.15	019.52	0
019	Calcium (%)	0049	2.2050	0.13000	2.1819	0.12488	0.04788	122	0.19	019.41	0
019	Calcium (%)	0885	2.2059	0.02760	2.1819	0.12488	0.04788	122	0.19	019.08	0
019	Calcium (%)	0881	2.2075	0.00500	2.1819	0.12488	0.04788	122	0.21	019.00	0
019	Calcium (%)	0512	2.2075	0.03500	2.1819	0.12488	0.04788	122	0.21	019.43	0
019	Calcium (%)	0849	2.2085	0.03700	2.1819	0.12488	0.04788	122	0.21	019.08	0
019	Calcium (%)	0683	2.2100	0.00000	2.1819	0.12488	0.04788	122	0.23	019.00	0
019	Calcium (%)	0187	2.2150	0.01000	2.1819	0.12488	0.04788	122	0.27	019.42	0
019	Calcium (%)	0918	2.2155	0.08700	2.1819	0.12488	0.04788	122	0.27	019.43	0
019	Calcium (%)	2195	2.2200	0.06000	2.1819	0.12488	0.04788	122	0.31	019.08	0
019	Calcium (%)	0035	2.2200	0.10000	2.1819	0.12488	0.04788	122	0.31	019.43	0
019	Calcium (%)	0895	2.2250	0.01000	2.1819	0.12488	0.04788	122	0.35	019.00	0
019	Calcium (%)	0612	2.2300	0.06000	2.1819	0.12488	0.04788	122	0.39	019.33	0
019	Calcium (%)	0629	2.2300	0.06000	2.1819	0.12488	0.04788	122	0.39	019.43	0
019	Calcium (%)	0032	2.2325	0.05900	2.1819	0.12488	0.04788	122	0.41	019.43	0
019	Calcium (%)	0186	2.2400	0.02000	2.1819	0.12488	0.04788	122	0.47	019.42	0
019	Calcium (%)	2089	2.2400	0.00000	2.1819	0.12488	0.04788	122	0.47	019.43	0
019	Calcium (%)	0100	2.2400	0.10000	2.1819	0.12488	0.04788	122	0.47	019.99	0
019	Calcium (%)	0014	2.2450	0.05000	2.1819	0.12488	0.04788	122	0.51	019.31	0
019	Calcium (%)	0035	2.2450	0.03000	2.1819	0.12488	0.04788	122	0.51	019.42	0
019	Calcium (%)	0043	2.2500	0.00000	2.1819	0.12488	0.04788	122	0.55	019.00	0
019	Calcium (%)	0413	2.2500	0.10000	2.1819	0.12488	0.04788	122	0.55	019.41	0
019	Calcium (%)	0186	2.2500	0.02000	2.1819	0.12488	0.04788	122	0.55	019.52	0
019	Calcium (%)	0572	2.2500	0.04000	2.1819	0.12488	0.04788	122	0.55	019.53	0
019	Calcium (%)	0006	2.2565	0.08900	2.1819	0.12488	0.04788	122	0.60	019.43	0
019	Calcium (%)	0897	2.2600	0.04000	2.1819	0.12488	0.04788	122	0.63	019.00	0
019	Calcium (%)	0036	2.2636	0.00390	2.1819	0.12488	0.04788	122	0.65	019.03	0
019	Calcium (%)	0689	2.2700	0.04000	2.1819	0.12488	0.04788	122	0.71	019.08	0
019	Calcium (%)	0358	2.2700	0.06000	2.1819	0.12488	0.04788	122	0.71	019.41	0
019	Calcium (%)	0083	2.2700	0.04000	2.1819	0.12488	0.04788	122	0.71	019.43	0
019	Calcium (%)	2053	2.2750	0.19000	2.1819	0.12488	0.04788	122	0.75	019.42	0
019	Calcium (%)	0874	2.2850	0.11000	2.1819	0.12488	0.04788	122	0.83	019.31	0
019	Calcium (%)	0598	2.2870	0.00200	2.1819	0.12488	0.04788	122	0.84	019.41	0
019	Calcium (%)	0520	2.3150	0.07000	2.1819	0.12488	0.04788	122	1.07	019.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	0142	2.3222	0.00730	2.1819	0.12488	0.04788	122	1.12	019.31	0
019	Calcium (%)	0868	2.3250	0.05000	2.1819	0.12488	0.04788	122	1.15	019.31	0
019	Calcium (%)	0353	2.3300	0.10000	2.1819	0.12488	0.04788	122	1.19	019.43	0
019	Calcium (%)	0160	2.3334	0.02720	2.1819	0.12488	0.04788	122	1.21	019.44	0
019	Calcium (%)	2006	2.3335	0.05700	2.1819	0.12488	0.04788	122	1.21	019.00	0
019	Calcium (%)	0108	2.3350	0.03000	2.1819	0.12488	0.04788	122	1.23	019.31	0
019	Calcium (%)	0357	2.3469	0.06290	2.1819	0.12488	0.04788	122	1.32	019.42	0
019	Calcium (%)	0027	2.3490	0.01400	2.1819	0.12488	0.04788	122	1.34	019.43	0
019	Calcium (%)	0687	2.3500	0.10000	2.1819	0.12488	0.04788	122	1.35	019.31	0
019	Calcium (%)	2006	2.3525	0.05900	2.1819	0.12488	0.04788	122	1.37	019.09	0
019	Calcium (%)	2126	2.3550	0.01000	2.1819	0.12488	0.04788	122	1.39	019.99	0
019	Calcium (%)	0226	2.4000	0.02000	2.1819	0.12488	0.04788	122	1.75	019.41	0
019	Calcium (%)	0300	2.4100	0.08000	2.1819	0.12488	0.04788	122	1.83	019.33	0
019	Calcium (%)	0405	2.4550	0.01000	2.1819	0.12488	0.04788	122	2.19	019.41	0
019	Calcium (%)	0425	2.4550	0.09000	2.1819	0.12488	0.04788	122	2.19	019.43	0
019	Calcium (%)	0089	2.4650	0.01000	2.1819	0.12488	0.04788	122	2.27	019.43	0
019	Calcium (%)	0726	2.5165	0.02300	2.1819	0.12488	0.04788	122	2.68	019.42	0
019	Calcium (%)	0034	2.5320	0.17000	2.1819	0.12488	0.04788	122	2.80	019.43	0
019	Calcium (%)	0190	2.5800	0.04000	2.1819	0.12488	0.04788	122	3.19	019.42	0
019	Calcium (%)	0123	2.8700	0.00000	2.1819	0.12488	0.04788	122	5.51	019.41	0
019	Calcium (%)	0169	74.040	2.8000	2.1819	0.12488	0.04788	122	575.40	019.32	2
021	Cobalt (mg / kg (ppm))	0675	2.6200	0.08000	4.0709	1.1011	0.21905	20	-1.32	021.43	0
021	Cobalt (mg / kg (ppm))	0918	2.7642	0.16590	4.0709	1.1011	0.21905	20	-1.19	021.53	0
021	Cobalt (mg / kg (ppm))	0910	2.7750	0.37000	4.0709	1.1011	0.21905	20	-1.18	021.52	0
021	Cobalt (mg / kg (ppm))	0955	2.9250	0.19000	4.0709	1.1011	0.21905	20	-1.04	021.41	0
021	Cobalt (mg / kg (ppm))	0693	2.9895	0.44100	4.0709	1.1011	0.21905	20	-0.98	021.42	0
021	Cobalt (mg / kg (ppm))	0047	3.1800	0.02000	4.0709	1.1011	0.21905	20	-0.81	021.52	0
021	Cobalt (mg / kg (ppm))	0045	3.5800	0.24000	4.0709	1.1011	0.21905	20	-0.45	021.42	0
021	Cobalt (mg / kg (ppm))	0011	3.7855	0.15900	4.0709	1.1011	0.21905	20	-0.26	021.41	0
021	Cobalt (mg / kg (ppm))	0345	3.9000	0.02000	4.0709	1.1011	0.21905	20	-0.16	021.43	0
021	Cobalt (mg / kg (ppm))	0171	4.0500	0.02000	4.0709	1.1011	0.21905	20	-0.02	021.41	0
021	Cobalt (mg / kg (ppm))	2053	4.0850	0.19000	4.0709	1.1011	0.21905	20	0.01	021.42	0
021	Cobalt (mg / kg (ppm))	0169	4.3700	0.24000	4.0709	1.1011	0.21905	20	0.27	021.43	0
021	Cobalt (mg / kg (ppm))	0689	4.5500	0.10000	4.0709	1.1011	0.21905	20	0.44	021.31	0
021	Cobalt (mg / kg (ppm))	0510	4.7600	0.04000	4.0709	1.1011	0.21905	20	0.63	021.43	0
021	Cobalt (mg / kg (ppm))	0208	4.8150	0.15000	4.0709	1.1011	0.21905	20	0.68	021.31	0
021	Cobalt (mg / kg (ppm))	0508	4.8980	0.23410	4.0709	1.1011	0.21905	20	0.75	021.43	0
021	Cobalt (mg / kg (ppm))	0560	5.0400	0.36000	4.0709	1.1011	0.21905	20	0.88	021.42	0
021	Cobalt (mg / kg (ppm))	0098	5.1375	0.82900	4.0709	1.1011	0.21905	20	0.97	021.53	0
021	Cobalt (mg / kg (ppm))	0939	5.4700	0.46000	4.0709	1.1011	0.21905	20	1.27	021.31	0
021	Cobalt (mg / kg (ppm))	0032	5.7330	0.07200	4.0709	1.1011	0.21905	20	1.51	021.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
021	Cobalt (mg / kg (ppm))	0186	16.500	27.000	4.0709	1.1011	0.21905	20	11.29	021.52	2
022	Copper (mg / kg (ppm))	0016	81.150	8.3000	120.58	7.9229	4.0227	86	-4.98	022.52	0
022	Copper (mg / kg (ppm))	0169	82.750	0.10000	120.58	7.9229	4.0227	86	-4.77	022.43	0
022	Copper (mg / kg (ppm))	2066	92.600	1.0000	120.58	7.9229	4.0227	86	-3.53	022.53	0
022	Copper (mg / kg (ppm))	2145	95.775	1.7700	120.58	7.9229	4.0227	86	-3.13	022.41	0
022	Copper (mg / kg (ppm))	0878	107.00	8.0000	120.58	7.9229	4.0227	86	-1.71	022.41	0
022	Copper (mg / kg (ppm))	0037	107.05	2.3000	120.58	7.9229	4.0227	86	-1.71	022.42	0
022	Copper (mg / kg (ppm))	0263	107.89	0.31000	120.58	7.9229	4.0227	86	-1.60	022.42	0
022	Copper (mg / kg (ppm))	0043	109.50	7.0000	120.58	7.9229	4.0227	86	-1.40	022.43	0
022	Copper (mg / kg (ppm))	0297	111.00	14.000	120.58	7.9229	4.0227	86	-1.21	022.43	0
022	Copper (mg / kg (ppm))	0034	111.30	4.2000	120.58	7.9229	4.0227	86	-1.17	022.43	0
022	Copper (mg / kg (ppm))	0910	112.00	6.0000	120.58	7.9229	4.0227	86	-1.08	022.41	0
022	Copper (mg / kg (ppm))	0049	112.92	7.5500	120.58	7.9229	4.0227	86	-0.97	022.41	0
022	Copper (mg / kg (ppm))	0870	113.10	4.4000	120.58	7.9229	4.0227	86	-0.94	022.42	0
022	Copper (mg / kg (ppm))	0955	113.50	7.0000	120.58	7.9229	4.0227	86	-0.89	022.41	0
022	Copper (mg / kg (ppm))	0045	114.00	4.0000	120.58	7.9229	4.0227	86	-0.83	022.42	0
022	Copper (mg / kg (ppm))	0536	114.13	2.0500	120.58	7.9229	4.0227	86	-0.81	022.31	0
022	Copper (mg / kg (ppm))	0345	114.50	1.0000	120.58	7.9229	4.0227	86	-0.77	022.43	0
022	Copper (mg / kg (ppm))	0939	114.92	0.66000	120.58	7.9229	4.0227	86	-0.71	022.31	0
022	Copper (mg / kg (ppm))	0175	115.00	2.0000	120.58	7.9229	4.0227	86	-0.70	022.31	0
022	Copper (mg / kg (ppm))	0629	115.00	0.00000	120.58	7.9229	4.0227	86	-0.70	022.31	0
022	Copper (mg / kg (ppm))	2152	115.03	0.20000	120.58	7.9229	4.0227	86	-0.70	022.41	0
022	Copper (mg / kg (ppm))	0358	115.77	10.900	120.58	7.9229	4.0227	86	-0.61	022.41	0
022	Copper (mg / kg (ppm))	0918	115.91	1.1600	120.58	7.9229	4.0227	86	-0.59	022.43	0
022	Copper (mg / kg (ppm))	0598	116.25	0.30000	120.58	7.9229	4.0227	86	-0.55	022.41	0
022	Copper (mg / kg (ppm))	0353	116.50	5.0000	120.58	7.9229	4.0227	86	-0.51	022.43	0
022	Copper (mg / kg (ppm))	0011	116.72	0.17500	120.58	7.9229	4.0227	86	-0.49	022.41	0
022	Copper (mg / kg (ppm))	0354	116.82	3.4400	120.58	7.9229	4.0227	86	-0.47	022.31	0
022	Copper (mg / kg (ppm))	0001	116.85	0.70000	120.58	7.9229	4.0227	86	-0.47	022.31	0
022	Copper (mg / kg (ppm))	0208	116.90	2.4000	120.58	7.9229	4.0227	86	-0.46	022.41	0
022	Copper (mg / kg (ppm))	0629	117.00	10.000	120.58	7.9229	4.0227	86	-0.45	022.43	0
022	Copper (mg / kg (ppm))	0019	117.35	3.3000	120.58	7.9229	4.0227	86	-0.41	022.41	0
022	Copper (mg / kg (ppm))	0278	117.90	1.2000	120.58	7.9229	4.0227	86	-0.34	022.42	0
022	Copper (mg / kg (ppm))	2188	118.00	2.0000	120.58	7.9229	4.0227	86	-0.33	022.31	0
022	Copper (mg / kg (ppm))	0407	118.04	0.67690	120.58	7.9229	4.0227	86	-0.32	022.41	0
022	Copper (mg / kg (ppm))	0098	118.65	5.1000	120.58	7.9229	4.0227	86	-0.24	022.43	0
022	Copper (mg / kg (ppm))	0171	119.00	2.0000	120.58	7.9229	4.0227	86	-0.20	022.41	0
022	Copper (mg / kg (ppm))	0405	119.05	1.9000	120.58	7.9229	4.0227	86	-0.19	022.41	0
022	Copper (mg / kg (ppm))	2089	119.24	12.140	120.58	7.9229	4.0227	86	-0.17	022.43	0
022	Copper (mg / kg (ppm))	0229	119.26	5.9300	120.58	7.9229	4.0227	86	-0.17	022.41	0
022	Copper (mg / kg (ppm))	0520	119.50	3.0000	120.58	7.9229	4.0227	86	-0.14	022.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
022	Copper (mg / kg (ppm))	0100	119.50	5.0000	120.58	7.9229	4.0227	86	-0.14	022.99	0
022	Copper (mg / kg (ppm))	0685	120.04	1.8400	120.58	7.9229	4.0227	86	-0.07	022.42	0
022	Copper (mg / kg (ppm))	0726	120.50	0.80000	120.58	7.9229	4.0227	86	-0.01	022.42	0
022	Copper (mg / kg (ppm))	0968	120.50	1.0000	120.58	7.9229	4.0227	86	-0.01	022.43	0
022	Copper (mg / kg (ppm))	0572	120.50	3.0000	120.58	7.9229	4.0227	86	-0.01	022.53	0
022	Copper (mg / kg (ppm))	0202	120.78	0.15000	120.58	7.9229	4.0227	86	0.03	022.43	0
022	Copper (mg / kg (ppm))	0508	120.79	8.0800	120.58	7.9229	4.0227	86	0.03	022.43	0
022	Copper (mg / kg (ppm))	0512	120.85	6.7000	120.58	7.9229	4.0227	86	0.03	022.43	0
022	Copper (mg / kg (ppm))	0723	121.00	0.00000	120.58	7.9229	4.0227	86	0.05	022.31	0
022	Copper (mg / kg (ppm))	0682	121.00	0.00000	120.58	7.9229	4.0227	86	0.05	022.43	0
022	Copper (mg / kg (ppm))	0047	121.85	8.1000	120.58	7.9229	4.0227	86	0.16	022.52	0
022	Copper (mg / kg (ppm))	0208	122.00	6.0000	120.58	7.9229	4.0227	86	0.18	022.31	0
022	Copper (mg / kg (ppm))	0357	122.48	10.429	120.58	7.9229	4.0227	86	0.24	022.42	0
022	Copper (mg / kg (ppm))	0014	122.50	7.0000	120.58	7.9229	4.0227	86	0.24	022.31	0
022	Copper (mg / kg (ppm))	0003	122.50	3.0000	120.58	7.9229	4.0227	86	0.24	022.41	0
022	Copper (mg / kg (ppm))	0098	122.50	2.6000	120.58	7.9229	4.0227	86	0.24	022.41	0
022	Copper (mg / kg (ppm))	0590	122.50	1.0000	120.58	7.9229	4.0227	86	0.24	022.99	0
022	Copper (mg / kg (ppm))	0693	122.78	6.1970	120.58	7.9229	4.0227	86	0.28	022.42	0
022	Copper (mg / kg (ppm))	0190	123.59	2.1200	120.58	7.9229	4.0227	86	0.38	022.42	0
022	Copper (mg / kg (ppm))	0425	123.60	1.2000	120.58	7.9229	4.0227	86	0.38	022.43	0
022	Copper (mg / kg (ppm))	0675	124.00	11.110	120.58	7.9229	4.0227	86	0.43	022.43	0
022	Copper (mg / kg (ppm))	0510	124.00	2.0000	120.58	7.9229	4.0227	86	0.43	022.43	0
022	Copper (mg / kg (ppm))	0948	124.77	0.68900	120.58	7.9229	4.0227	86	0.53	022.33	0
022	Copper (mg / kg (ppm))	0186	125.00	2.0000	120.58	7.9229	4.0227	86	0.56	022.52	0
022	Copper (mg / kg (ppm))	0242	125.00	6.0000	120.58	7.9229	4.0227	86	0.56	022.99	0
022	Copper (mg / kg (ppm))	0205	125.50	5.0000	120.58	7.9229	4.0227	86	0.62	022.42	0
022	Copper (mg / kg (ppm))	0868	126.52	0.63000	120.58	7.9229	4.0227	86	0.75	022.31	0
022	Copper (mg / kg (ppm))	0010	127.00	2.0000	120.58	7.9229	4.0227	86	0.81	022.33	0
022	Copper (mg / kg (ppm))	0083	127.50	3.0000	120.58	7.9229	4.0227	86	0.87	022.43	0
022	Copper (mg / kg (ppm))	0511	128.00	2.0000	120.58	7.9229	4.0227	86	0.94	022.41	0
022	Copper (mg / kg (ppm))	0032	128.25	11.100	120.58	7.9229	4.0227	86	0.97	022.43	0
022	Copper (mg / kg (ppm))	0413	128.50	7.0000	120.58	7.9229	4.0227	86	1.00	022.42	0
022	Copper (mg / kg (ppm))	0560	128.75	3.5000	120.58	7.9229	4.0227	86	1.03	022.42	0
022	Copper (mg / kg (ppm))	2196	129.00	0.00000	120.58	7.9229	4.0227	86	1.06	022.31	0
022	Copper (mg / kg (ppm))	0187	129.41	0.99000	120.58	7.9229	4.0227	86	1.11	022.42	0
022	Copper (mg / kg (ppm))	0294	130.63	3.3900	120.58	7.9229	4.0227	86	1.27	022.42	0
022	Copper (mg / kg (ppm))	0042	133.50	11.000	120.58	7.9229	4.0227	86	1.63	022.42	0
022	Copper (mg / kg (ppm))	0027	134.51	13.869	120.58	7.9229	4.0227	86	1.76	022.43	0
022	Copper (mg / kg (ppm))	2195	134.79	1.8000	120.58	7.9229	4.0227	86	1.79	022.31	0
022	Copper (mg / kg (ppm))	0226	136.50	3.0000	120.58	7.9229	4.0227	86	2.01	022.41	0
022	Copper (mg / kg (ppm))	2053	136.50	3.0000	120.58	7.9229	4.0227	86	2.01	022.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
022	Copper (mg / kg (ppm))	0035	138.00	8.0000	120.58	7.9229	4.0227	86	2.20	022.43	0
022	Copper (mg / kg (ppm))	0035	143.00	8.0000	120.58	7.9229	4.0227	86	2.83	022.42	0
022	Copper (mg / kg (ppm))	0186	145.50	7.0000	120.58	7.9229	4.0227	86	3.15	022.42	0
022	Copper (mg / kg (ppm))	0689	148.15	2.5000	120.58	7.9229	4.0227	86	3.48	022.31	0
022	Copper (mg / kg (ppm))	0609	165.00	0.00000	120.58	7.9229	4.0227	86	5.61	022.32	0
022	Copper (mg / kg (ppm))	0692	129.50	33.000	120.58	7.9229	4.0227	86	1.13	022.42	1
022	Copper (mg / kg (ppm))	2178	191.50	63.000	120.58	7.9229	4.0227	86	8.95	022.33	1
023	Fluorine (mg / kg (ppm))	0910	28.900	0.00000				1		023.01	0
025	Iron (mg / kg (ppm))	0536	29.325	5.2500	770.01	80.339	18.378	72	-9.22	025.31	0
025	Iron (mg / kg (ppm))	0726	285.30	6.4000	770.01	80.339	18.378	72	-6.03	025.42	0
025	Iron (mg / kg (ppm))	0187	354.79	16.370	770.01	80.339	18.378	72	-5.17	025.42	0
025	Iron (mg / kg (ppm))	2066	449.50	27.000	770.01	80.339	18.378	72	-3.99	025.53	0
025	Iron (mg / kg (ppm))	0190	480.72	35.600	770.01	80.339	18.378	72	-3.60	025.42	0
025	Iron (mg / kg (ppm))	0169	520.00	2.0000	770.01	80.339	18.378	72	-3.11	025.43	0
025	Iron (mg / kg (ppm))	0047	559.95	3.5000	770.01	80.339	18.378	72	-2.61	025.52	0
025	Iron (mg / kg (ppm))	0609	627.00	8.0000	770.01	80.339	18.378	72	-1.78	025.32	0
025	Iron (mg / kg (ppm))	0278	636.35	52.700	770.01	80.339	18.378	72	-1.66	025.42	0
025	Iron (mg / kg (ppm))	0598	646.70	9.0000	770.01	80.339	18.378	72	-1.53	025.41	0
025	Iron (mg / kg (ppm))	0560	658.65	7.5000	770.01	80.339	18.378	72	-1.39	025.42	0
025	Iron (mg / kg (ppm))	0968	671.50	9.0000	770.01	80.339	18.378	72	-1.23	025.43	0
025	Iron (mg / kg (ppm))	0345	691.00	2.0000	770.01	80.339	18.378	72	-0.98	025.43	0
025	Iron (mg / kg (ppm))	0692	694.50	35.000	770.01	80.339	18.378	72	-0.94	025.42	0
025	Iron (mg / kg (ppm))	0294	700.25	0.10000	770.01	80.339	18.378	72	-0.87	025.42	0
025	Iron (mg / kg (ppm))	0955	709.00	82.000	770.01	80.339	18.378	72	-0.76	025.41	0
025	Iron (mg / kg (ppm))	2145	711.00	5.8000	770.01	80.339	18.378	72	-0.73	025.41	0
025	Iron (mg / kg (ppm))	0874	721.00	12.000	770.01	80.339	18.378	72	-0.61	025.31	0
025	Iron (mg / kg (ppm))	0154	725.54	13.670	770.01	80.339	18.378	72	-0.55	025.52	0
025	Iron (mg / kg (ppm))	0353	736.00	20.000	770.01	80.339	18.378	72	-0.42	025.43	0
025	Iron (mg / kg (ppm))	0675	739.97	2.5500	770.01	80.339	18.378	72	-0.37	025.43	0
025	Iron (mg / kg (ppm))	2089	742.52	32.710	770.01	80.339	18.378	72	-0.34	025.43	0
025	Iron (mg / kg (ppm))	2192	743.00	38.000	770.01	80.339	18.378	72	-0.34	025.43	0
025	Iron (mg / kg (ppm))	0045	748.50	15.000	770.01	80.339	18.378	72	-0.27	025.42	0
025	Iron (mg / kg (ppm))	0175	753.00	18.000	770.01	80.339	18.378	72	-0.21	025.31	0
025	Iron (mg / kg (ppm))	0037	754.30	4.2000	770.01	80.339	18.378	72	-0.20	025.42	0
025	Iron (mg / kg (ppm))	0407	755.87	2.5516	770.01	80.339	18.378	72	-0.18	025.41	0
025	Iron (mg / kg (ppm))	0629	756.50	11.000	770.01	80.339	18.378	72	-0.17	025.43	0
025	Iron (mg / kg (ppm))	0870	760.00	51.000	770.01	80.339	18.378	72	-0.12	025.42	0
025	Iron (mg / kg (ppm))	2152	761.92	1.0800	770.01	80.339	18.378	72	-0.10	025.41	0
025	Iron (mg / kg (ppm))	0918	764.42	47.770	770.01	80.339	18.378	72	-0.07	025.43	0
025	Iron (mg / kg (ppm))	0413	768.00	6.0000	770.01	80.339	18.378	72	-0.03	025.42	0
025	Iron (mg / kg (ppm))	0003	775.00	12.000	770.01	80.339	18.378	72	0.06	025.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
025	Iron (mg / kg (ppm))	0878	775.00	22.000	770.01	80.339	18.378	72	0.06	025.41	0
025	Iron (mg / kg (ppm))	0693	776.59	50.470	770.01	80.339	18.378	72	0.08	025.42	0
025	Iron (mg / kg (ppm))	0939	780.45	0.28000	770.01	80.339	18.378	72	0.13	025.31	0
025	Iron (mg / kg (ppm))	0100	781.00	42.000	770.01	80.339	18.378	72	0.14	025.99	0
025	Iron (mg / kg (ppm))	0034	782.70	8.0000	770.01	80.339	18.378	72	0.16	025.43	0
025	Iron (mg / kg (ppm))	0520	784.50	25.000	770.01	80.339	18.378	72	0.18	025.41	0
025	Iron (mg / kg (ppm))	0511	788.50	7.0000	770.01	80.339	18.378	72	0.23	025.41	0
025	Iron (mg / kg (ppm))	0171	789.50	9.0000	770.01	80.339	18.378	72	0.24	025.41	0
025	Iron (mg / kg (ppm))	0098	789.75	9.3000	770.01	80.339	18.378	72	0.25	025.41	0
025	Iron (mg / kg (ppm))	0425	789.80	8.8000	770.01	80.339	18.378	72	0.25	025.43	0
025	Iron (mg / kg (ppm))	0208	792.10	18.800	770.01	80.339	18.378	72	0.27	025.41	0
025	Iron (mg / kg (ppm))	0229	797.30	0.53000	770.01	80.339	18.378	72	0.34	025.41	0
025	Iron (mg / kg (ppm))	0032	803.65	22.300	770.01	80.339	18.378	72	0.42	025.43	0
025	Iron (mg / kg (ppm))	2195	805.27	12.220	770.01	80.339	18.378	72	0.44	025.31	0
025	Iron (mg / kg (ppm))	0098	805.65	19.500	770.01	80.339	18.378	72	0.44	025.43	0
025	Iron (mg / kg (ppm))	0011	806.20	0.50500	770.01	80.339	18.378	72	0.45	025.41	0
025	Iron (mg / kg (ppm))	0510	809.00	26.000	770.01	80.339	18.378	72	0.49	025.43	0
025	Iron (mg / kg (ppm))	0910	812.50	45.000	770.01	80.339	18.378	72	0.53	025.41	0
025	Iron (mg / kg (ppm))	0226	819.50	5.0000	770.01	80.339	18.378	72	0.62	025.41	0
025	Iron (mg / kg (ppm))	0049	821.18	3.2000	770.01	80.339	18.378	72	0.64	025.41	0
025	Iron (mg / kg (ppm))	0689	823.40	31.000	770.01	80.339	18.378	72	0.66	025.31	0
025	Iron (mg / kg (ppm))	0242	827.00	26.000	770.01	80.339	18.378	72	0.71	025.99	0
025	Iron (mg / kg (ppm))	0682	828.49	0.00000	770.01	80.339	18.378	72	0.73	025.43	0
025	Iron (mg / kg (ppm))	0019	831.30	2.6000	770.01	80.339	18.378	72	0.76	025.41	0
025	Iron (mg / kg (ppm))	2196	832.00	0.00000	770.01	80.339	18.378	72	0.77	025.31	0
025	Iron (mg / kg (ppm))	0358	832.63	0.29000	770.01	80.339	18.378	72	0.78	025.41	0
025	Iron (mg / kg (ppm))	0083	839.50	29.000	770.01	80.339	18.378	72	0.86	025.43	0
025	Iron (mg / kg (ppm))	0354	841.08	3.7100	770.01	80.339	18.378	72	0.88	025.31	0
025	Iron (mg / kg (ppm))	0868	841.50	3.4600	770.01	80.339	18.378	72	0.89	025.31	0
025	Iron (mg / kg (ppm))	0263	844.38	1.8500	770.01	80.339	18.378	72	0.93	025.42	0
025	Iron (mg / kg (ppm))	0208	847.50	1.0000	770.01	80.339	18.378	72	0.96	025.31	0
025	Iron (mg / kg (ppm))	0297	850.00	124.00	770.01	80.339	18.378	72	1.00	025.43	0
025	Iron (mg / kg (ppm))	2188	856.00	20.000	770.01	80.339	18.378	72	1.07	025.31	0
025	Iron (mg / kg (ppm))	0001	869.55	4.7000	770.01	80.339	18.378	72	1.24	025.31	0
025	Iron (mg / kg (ppm))	0512	873.35	47.900	770.01	80.339	18.378	72	1.29	025.43	0
025	Iron (mg / kg (ppm))	0723	885.00	0.00000	770.01	80.339	18.378	72	1.43	025.31	0
025	Iron (mg / kg (ppm))	2178	911.00	96.000	770.01	80.339	18.378	72	1.75	025.33	0
025	Iron (mg / kg (ppm))	0405	1,031.0	2.0000	770.01	80.339	18.378	72	3.25	025.41	0
025	Iron (mg / kg (ppm))	0948	1,086.1	0.03100	770.01	80.339	18.378	72	3.93	025.33	0
025	Iron (mg / kg (ppm))	0042	599.00	220.00	770.01	80.339	18.378	72	-2.13	025.42	1
025	Iron (mg / kg (ppm))	0160	640.00	146.00	770.01	80.339	18.378	72	-1.62	025.51	1

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
025	Iron (mg / kg (ppm))	0508	881.82	150.40	770.01	80.339	18.378	72	1.39	025.43	1
027	Magnesium (%)	0536	0.40500	0.01000	0.93399	0.05299	0.02442	90	-9.98	027.31	0
027	Magnesium (%)	0889	0.68385	0.05550	0.93399	0.05299	0.02442	90	-4.72	027.99	0
027	Magnesium (%)	2066	0.73000	0.06000	0.93399	0.05299	0.02442	90	-3.85	027.53	0
027	Magnesium (%)	2188	0.74000	0.00000	0.93399	0.05299	0.02442	90	-3.66	027.31	0
027	Magnesium (%)	0939	0.84000	0.02000	0.93399	0.05299	0.02442	90	-1.77	027.31	0
027	Magnesium (%)	0511	0.86000	0.08000	0.93399	0.05299	0.02442	90	-1.40	027.41	0
027	Magnesium (%)	0297	0.86000	0.02000	0.93399	0.05299	0.02442	90	-1.40	027.43	0
027	Magnesium (%)	0171	0.86600	0.00600	0.93399	0.05299	0.02442	90	-1.28	027.41	0
027	Magnesium (%)	0294	0.87000	0.02000	0.93399	0.05299	0.02442	90	-1.21	027.42	0
027	Magnesium (%)	0045	0.87150	0.08900	0.93399	0.05299	0.02442	90	-1.18	027.42	0
027	Magnesium (%)	0032	0.87460	0.02960	0.93399	0.05299	0.02442	90	-1.12	027.43	0
027	Magnesium (%)	2089	0.87500	0.01000	0.93399	0.05299	0.02442	90	-1.11	027.43	0
027	Magnesium (%)	0870	0.87645	0.04650	0.93399	0.05299	0.02442	90	-1.09	027.42	0
027	Magnesium (%)	0042	0.87750	0.03100	0.93399	0.05299	0.02442	90	-1.07	027.42	0
027	Magnesium (%)	0693	0.88450	0.02900	0.93399	0.05299	0.02442	90	-0.93	027.42	0
027	Magnesium (%)	0098	0.88500	0.01000	0.93399	0.05299	0.02442	90	-0.92	027.43	0
027	Magnesium (%)	0098	0.89000	0.02000	0.93399	0.05299	0.02442	90	-0.83	027.41	0
027	Magnesium (%)	0175	0.89500	0.01000	0.93399	0.05299	0.02442	90	-0.74	027.31	0
027	Magnesium (%)	0049	0.89500	0.09000	0.93399	0.05299	0.02442	90	-0.74	027.41	0
027	Magnesium (%)	0345	0.89750	0.00700	0.93399	0.05299	0.02442	90	-0.69	027.43	0
027	Magnesium (%)	0354	0.90000	0.00000	0.93399	0.05299	0.02442	90	-0.64	027.31	0
027	Magnesium (%)	0037	0.90000	0.02000	0.93399	0.05299	0.02442	90	-0.64	027.42	0
027	Magnesium (%)	0510	0.90000	0.00000	0.93399	0.05299	0.02442	90	-0.64	027.43	0
027	Magnesium (%)	0918	0.90250	0.02500	0.93399	0.05299	0.02442	90	-0.59	027.43	0
027	Magnesium (%)	0910	0.90500	0.01000	0.93399	0.05299	0.02442	90	-0.55	027.41	0
027	Magnesium (%)	0154	0.90820	0.00040	0.93399	0.05299	0.02442	90	-0.49	027.52	0
027	Magnesium (%)	0598	0.90840	0.01000	0.93399	0.05299	0.02442	90	-0.48	027.41	0
027	Magnesium (%)	0868	0.91000	0.00000	0.93399	0.05299	0.02442	90	-0.45	027.31	0
027	Magnesium (%)	0011	0.91260	0.03400	0.93399	0.05299	0.02442	90	-0.40	027.41	0
027	Magnesium (%)	0263	0.91275	0.00070	0.93399	0.05299	0.02442	90	-0.40	027.42	0
027	Magnesium (%)	0353	0.91500	0.01000	0.93399	0.05299	0.02442	90	-0.36	027.43	0
027	Magnesium (%)	0034	0.91595	0.00670	0.93399	0.05299	0.02442	90	-0.34	027.43	0
027	Magnesium (%)	0560	0.91600	0.01040	0.93399	0.05299	0.02442	90	-0.34	027.42	0
027	Magnesium (%)	0208	0.91650	0.02500	0.93399	0.05299	0.02442	90	-0.33	027.31	0
027	Magnesium (%)	0139	0.91920	0.01100	0.93399	0.05299	0.02442	90	-0.28	027.41	0
027	Magnesium (%)	0187	0.92130	0.00120	0.93399	0.05299	0.02442	90	-0.24	027.42	0
027	Magnesium (%)	0014	0.92200	0.01000	0.93399	0.05299	0.02442	90	-0.23	027.31	0
027	Magnesium (%)	0010	0.92500	0.01000	0.93399	0.05299	0.02442	90	-0.17	027.33	0
027	Magnesium (%)	0278	0.92500	0.01000	0.93399	0.05299	0.02442	90	-0.17	027.42	0
027	Magnesium (%)	0035	0.92500	0.03000	0.93399	0.05299	0.02442	90	-0.17	027.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
027	Magnesium (%)	0590	0.92500	0.01000	0.93399	0.05299	0.02442	90	-0.17	027.99	0
027	Magnesium (%)	0692	0.93000	0.20000	0.93399	0.05299	0.02442	90	-0.08	027.42	0
027	Magnesium (%)	0003	0.93000	0.00000	0.93399	0.05299	0.02442	90	-0.08	027.41	0
027	Magnesium (%)	2145	0.93000	0.00000	0.93399	0.05299	0.02442	90	-0.08	027.41	0
027	Magnesium (%)	0682	0.93000	0.00000	0.93399	0.05299	0.02442	90	-0.08	027.43	0
027	Magnesium (%)	0955	0.93200	0.01200	0.93399	0.05299	0.02442	90	-0.04	027.44	0
027	Magnesium (%)	0229	0.93300	0.01800	0.93399	0.05299	0.02442	90	-0.02	027.41	0
027	Magnesium (%)	0208	0.93400	0.01200	0.93399	0.05299	0.02442	90	0.00	027.41	0
027	Magnesium (%)	0006	0.93400	0.01400	0.93399	0.05299	0.02442	90	0.00	027.43	0
027	Magnesium (%)	0878	0.93500	0.01000	0.93399	0.05299	0.02442	90	0.02	027.41	0
027	Magnesium (%)	0685	0.93500	0.01000	0.93399	0.05299	0.02442	90	0.02	027.42	0
027	Magnesium (%)	0242	0.93500	0.01000	0.93399	0.05299	0.02442	90	0.02	027.99	0
027	Magnesium (%)	0512	0.93790	0.02560	0.93399	0.05299	0.02442	90	0.07	027.43	0
027	Magnesium (%)	0036	0.93850	0.00580	0.93399	0.05299	0.02442	90	0.09	027.43	0
027	Magnesium (%)	0357	0.94105	0.06010	0.93399	0.05299	0.02442	90	0.13	027.42	0
027	Magnesium (%)	0001	0.94320	0.00680	0.93399	0.05299	0.02442	90	0.17	027.31	0
027	Magnesium (%)	0142	0.94485	0.01130	0.93399	0.05299	0.02442	90	0.20	027.31	0
027	Magnesium (%)	0609	0.94500	0.01000	0.93399	0.05299	0.02442	90	0.21	027.32	0
027	Magnesium (%)	0019	0.94500	0.03000	0.93399	0.05299	0.02442	90	0.21	027.41	0
027	Magnesium (%)	0413	0.94500	0.03000	0.93399	0.05299	0.02442	90	0.21	027.41	0
027	Magnesium (%)	0202	0.94550	0.01700	0.93399	0.05299	0.02442	90	0.22	027.43	0
027	Magnesium (%)	0508	0.94910	0.15520	0.93399	0.05299	0.02442	90	0.29	027.43	0
027	Magnesium (%)	0629	0.95000	0.02000	0.93399	0.05299	0.02442	90	0.30	027.31	0
027	Magnesium (%)	0047	0.95000	0.04000	0.93399	0.05299	0.02442	90	0.30	027.52	0
027	Magnesium (%)	0948	0.96050	0.00900	0.93399	0.05299	0.02442	90	0.50	027.33	0
027	Magnesium (%)	0043	0.96050	0.00100	0.93399	0.05299	0.02442	90	0.50	027.43	0
027	Magnesium (%)	0675	0.96500	0.01000	0.93399	0.05299	0.02442	90	0.59	027.43	0
027	Magnesium (%)	0186	0.96950	0.00500	0.93399	0.05299	0.02442	90	0.67	027.52	0
027	Magnesium (%)	0520	0.97000	0.06000	0.93399	0.05299	0.02442	90	0.68	027.41	0
027	Magnesium (%)	0027	0.97000	0.01400	0.93399	0.05299	0.02442	90	0.68	027.43	0
027	Magnesium (%)	0100	0.97000	0.08000	0.93399	0.05299	0.02442	90	0.68	027.99	0
027	Magnesium (%)	0035	0.98000	0.02000	0.93399	0.05299	0.02442	90	0.87	027.42	0
027	Magnesium (%)	0083	0.98000	0.02000	0.93399	0.05299	0.02442	90	0.87	027.43	0
027	Magnesium (%)	0186	0.98250	0.02100	0.93399	0.05299	0.02442	90	0.92	027.42	0
027	Magnesium (%)	0968	0.98400	0.00400	0.93399	0.05299	0.02442	90	0.94	027.43	0
027	Magnesium (%)	0160	0.98680	0.03200	0.93399	0.05299	0.02442	90	1.00	027.44	0
027	Magnesium (%)	0572	0.98700	0.00400	0.93399	0.05299	0.02442	90	1.00	027.53	0
027	Magnesium (%)	0874	1.0075	0.04500	0.93399	0.05299	0.02442	90	1.39	027.31	0
027	Magnesium (%)	0629	1.0100	0.02000	0.93399	0.05299	0.02442	90	1.43	027.43	0
027	Magnesium (%)	0407	1.0120	0.03200	0.93399	0.05299	0.02442	90	1.47	027.41	0
027	Magnesium (%)	0612	1.0150	0.01000	0.93399	0.05299	0.02442	90	1.53	027.33	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
027	Magnesium (%)	0358	1.0250	0.01000	0.93399	0.05299	0.02442	90	1.72	027.41	0
027	Magnesium (%)	0169	1.0300	0.12000	0.93399	0.05299	0.02442	90	1.81	027.32	0
027	Magnesium (%)	0226	1.0300	0.00000	0.93399	0.05299	0.02442	90	1.81	027.41	0
027	Magnesium (%)	0190	1.0450	0.03000	0.93399	0.05299	0.02442	90	2.09	027.42	0
027	Magnesium (%)	0089	1.0500	0.00000	0.93399	0.05299	0.02442	90	2.19	027.43	0
027	Magnesium (%)	0425	1.0500	0.04000	0.93399	0.05299	0.02442	90	2.19	027.43	0
027	Magnesium (%)	0405	1.0660	0.00800	0.93399	0.05299	0.02442	90	2.49	027.41	0
027	Magnesium (%)	0726	1.0775	0.00700	0.93399	0.05299	0.02442	90	2.71	027.42	0
027	Magnesium (%)	0689	1.2300	0.02000	0.93399	0.05299	0.02442	90	5.59	027.31	0
027	Magnesium (%)	2195	9,931.0	48.000	0.93399	0.05299	0.02442	90	187392.71	027.31	2
028	Manganese (mg / kg (ppm))	0948	209.01	0.25300	292.33	21.016	7.3359	83	-3.96	028.33	0
028	Manganese (mg / kg (ppm))	2066	232.00	24.000	292.33	21.016	7.3359	83	-2.87	028.53	0
028	Manganese (mg / kg (ppm))	0692	237.00	14.000	292.33	21.016	7.3359	83	-2.63	028.42	0
028	Manganese (mg / kg (ppm))	0685	251.66	0.05000	292.33	21.016	7.3359	83	-1.94	028.42	0
028	Manganese (mg / kg (ppm))	0353	253.00	10.000	292.33	21.016	7.3359	83	-1.87	028.43	0
028	Manganese (mg / kg (ppm))	2145	257.20	1.8000	292.33	21.016	7.3359	83	-1.67	028.41	0
028	Manganese (mg / kg (ppm))	2188	259.00	8.0000	292.33	21.016	7.3359	83	-1.59	028.31	0
028	Manganese (mg / kg (ppm))	0955	261.00	12.000	292.33	21.016	7.3359	83	-1.49	028.41	0
028	Manganese (mg / kg (ppm))	2192	265.58	32.840	292.33	21.016	7.3359	83	-1.27	028.53	0
028	Manganese (mg / kg (ppm))	0878	266.50	7.0000	292.33	21.016	7.3359	83	-1.23	028.41	0
028	Manganese (mg / kg (ppm))	0043	267.50	3.0000	292.33	21.016	7.3359	83	-1.18	028.43	0
028	Manganese (mg / kg (ppm))	0278	271.00	2.0000	292.33	21.016	7.3359	83	-1.02	028.42	0
028	Manganese (mg / kg (ppm))	0693	271.66	0.93800	292.33	21.016	7.3359	83	-0.98	028.42	0
028	Manganese (mg / kg (ppm))	0629	273.50	1.0000	292.33	21.016	7.3359	83	-0.90	028.31	0
028	Manganese (mg / kg (ppm))	0918	274.22	0.54000	292.33	21.016	7.3359	83	-0.86	028.43	0
028	Manganese (mg / kg (ppm))	0037	274.55	1.7000	292.33	21.016	7.3359	83	-0.85	028.42	0
028	Manganese (mg / kg (ppm))	0675	274.97	3.0600	292.33	21.016	7.3359	83	-0.83	028.43	0
028	Manganese (mg / kg (ppm))	2089	275.10	13.850	292.33	21.016	7.3359	83	-0.82	028.43	0
028	Manganese (mg / kg (ppm))	0226	275.50	1.0000	292.33	21.016	7.3359	83	-0.80	028.41	0
028	Manganese (mg / kg (ppm))	0263	276.40	2.3300	292.33	21.016	7.3359	83	-0.76	028.42	0
028	Manganese (mg / kg (ppm))	0407	276.40	0.17610	292.33	21.016	7.3359	83	-0.76	028.41	0
028	Manganese (mg / kg (ppm))	0045	277.00	10.000	292.33	21.016	7.3359	83	-0.73	028.42	0
028	Manganese (mg / kg (ppm))	0510	278.00	6.0000	292.33	21.016	7.3359	83	-0.68	028.43	0
028	Manganese (mg / kg (ppm))	0520	278.50	9.0000	292.33	21.016	7.3359	83	-0.66	028.41	0
028	Manganese (mg / kg (ppm))	0049	278.75	16.890	292.33	21.016	7.3359	83	-0.65	028.41	0
028	Manganese (mg / kg (ppm))	0689	278.95	12.300	292.33	21.016	7.3359	83	-0.64	028.31	0
028	Manganese (mg / kg (ppm))	0171	280.00	2.0000	292.33	21.016	7.3359	83	-0.59	028.41	0
028	Manganese (mg / kg (ppm))	0870	280.55	2.9000	292.33	21.016	7.3359	83	-0.56	028.42	0
028	Manganese (mg / kg (ppm))	0511	281.50	3.0000	292.33	21.016	7.3359	83	-0.52	028.41	0
028	Manganese (mg / kg (ppm))	0034	283.50	9.6000	292.33	21.016	7.3359	83	-0.42	028.43	0
028	Manganese (mg / kg (ppm))	0345	284.00	6.0000	292.33	21.016	7.3359	83	-0.40	028.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
028	Manganese (mg / kg (ppm))	0011	288.78	0.72700	292.33	21.016	7.3359	83	-0.17	028.41	0
028	Manganese (mg / kg (ppm))	0098	290.00	2.2000	292.33	21.016	7.3359	83	-0.11	028.41	0
028	Manganese (mg / kg (ppm))	2053	290.00	12.000	292.33	21.016	7.3359	83	-0.11	028.42	0
028	Manganese (mg / kg (ppm))	0297	290.00	8.0000	292.33	21.016	7.3359	83	-0.11	028.43	0
028	Manganese (mg / kg (ppm))	0175	291.00	18.000	292.33	21.016	7.3359	83	-0.06	028.31	0
028	Manganese (mg / kg (ppm))	0968	293.00	2.0000	292.33	21.016	7.3359	83	0.03	028.43	0
028	Manganese (mg / kg (ppm))	0536	293.98	6.2500	292.33	21.016	7.3359	83	0.08	028.31	0
028	Manganese (mg / kg (ppm))	0208	294.45	6.1000	292.33	21.016	7.3359	83	0.10	028.41	0
028	Manganese (mg / kg (ppm))	0598	294.60	6.2000	292.33	21.016	7.3359	83	0.11	028.41	0
028	Manganese (mg / kg (ppm))	0425	295.25	4.9000	292.33	21.016	7.3359	83	0.14	028.43	0
028	Manganese (mg / kg (ppm))	0001	295.35	13.300	292.33	21.016	7.3359	83	0.14	028.31	0
028	Manganese (mg / kg (ppm))	0042	296.00	4.0000	292.33	21.016	7.3359	83	0.17	028.42	0
028	Manganese (mg / kg (ppm))	0098	296.65	0.50000	292.33	21.016	7.3359	83	0.21	028.43	0
028	Manganese (mg / kg (ppm))	0413	297.00	8.0000	292.33	21.016	7.3359	83	0.22	028.42	0
028	Manganese (mg / kg (ppm))	0572	297.00	10.000	292.33	21.016	7.3359	83	0.22	028.53	0
028	Manganese (mg / kg (ppm))	0590	297.00	12.000	292.33	21.016	7.3359	83	0.22	028.99	0
028	Manganese (mg / kg (ppm))	0187	297.07	0.06000	292.33	21.016	7.3359	83	0.23	028.42	0
028	Manganese (mg / kg (ppm))	0560	297.45	7.9000	292.33	21.016	7.3359	83	0.24	028.42	0
028	Manganese (mg / kg (ppm))	2196	298.00	0.00000	292.33	21.016	7.3359	83	0.27	028.31	0
028	Manganese (mg / kg (ppm))	2195	298.26	2.4400	292.33	21.016	7.3359	83	0.28	028.31	0
028	Manganese (mg / kg (ppm))	0508	298.88	29.470	292.33	21.016	7.3359	83	0.31	028.43	0
028	Manganese (mg / kg (ppm))	0723	299.00	0.00000	292.33	21.016	7.3359	83	0.32	028.31	0
028	Manganese (mg / kg (ppm))	0229	299.28	2.1400	292.33	21.016	7.3359	83	0.33	028.41	0
028	Manganese (mg / kg (ppm))	0208	300.50	21.000	292.33	21.016	7.3359	83	0.39	028.31	0
028	Manganese (mg / kg (ppm))	0019	302.10	10.200	292.33	21.016	7.3359	83	0.46	028.41	0
028	Manganese (mg / kg (ppm))	0512	302.50	1.4000	292.33	21.016	7.3359	83	0.48	028.43	0
028	Manganese (mg / kg (ppm))	0629	303.00	12.000	292.33	21.016	7.3359	83	0.51	028.43	0
028	Manganese (mg / kg (ppm))	0939	303.92	2.7900	292.33	21.016	7.3359	83	0.55	028.31	0
028	Manganese (mg / kg (ppm))	0190	303.96	3.6300	292.33	21.016	7.3359	83	0.55	028.42	0
028	Manganese (mg / kg (ppm))	0100	304.50	13.000	292.33	21.016	7.3359	83	0.58	028.99	0
028	Manganese (mg / kg (ppm))	0242	304.50	1.0000	292.33	21.016	7.3359	83	0.58	028.99	0
028	Manganese (mg / kg (ppm))	0682	304.99	0.00000	292.33	21.016	7.3359	83	0.60	028.43	0
028	Manganese (mg / kg (ppm))	0726	305.55	5.3000	292.33	21.016	7.3359	83	0.63	028.42	0
028	Manganese (mg / kg (ppm))	0032	306.05	3.3000	292.33	21.016	7.3359	83	0.65	028.43	0
028	Manganese (mg / kg (ppm))	0910	307.50	15.000	292.33	21.016	7.3359	83	0.72	028.41	0
028	Manganese (mg / kg (ppm))	0014	308.00	12.000	292.33	21.016	7.3359	83	0.75	028.31	0
028	Manganese (mg / kg (ppm))	0047	308.25	4.1000	292.33	21.016	7.3359	83	0.76	028.52	0
028	Manganese (mg / kg (ppm))	0003	311.00	4.0000	292.33	21.016	7.3359	83	0.89	028.41	0
028	Manganese (mg / kg (ppm))	0354	311.02	1.3000	292.33	21.016	7.3359	83	0.89	028.31	0
028	Manganese (mg / kg (ppm))	0358	313.27	19.950	292.33	21.016	7.3359	83	1.00	028.41	0
028	Manganese (mg / kg (ppm))	0868	313.33	2.8700	292.33	21.016	7.3359	83	1.00	028.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
028	Manganese (mg / kg (ppm))	0083	314.00	4.0000	292.33	21.016	7.3359	83	1.03	028.43	0
028	Manganese (mg / kg (ppm))	0186	316.00	0.00000	292.33	21.016	7.3359	83	1.13	028.52	0
028	Manganese (mg / kg (ppm))	0027	316.58	19.175	292.33	21.016	7.3359	83	1.15	028.43	0
028	Manganese (mg / kg (ppm))	0357	318.27	14.413	292.33	21.016	7.3359	83	1.23	028.42	0
028	Manganese (mg / kg (ppm))	0202	323.44	2.5000	292.33	21.016	7.3359	83	1.48	028.43	0
028	Manganese (mg / kg (ppm))	0035	327.00	6.0000	292.33	21.016	7.3359	83	1.65	028.42	0
028	Manganese (mg / kg (ppm))	0035	330.00	0.00000	292.33	21.016	7.3359	83	1.79	028.43	0
028	Manganese (mg / kg (ppm))	0186	333.50	11.000	292.33	21.016	7.3359	83	1.96	028.42	0
028	Manganese (mg / kg (ppm))	0609	340.00	10.000	292.33	21.016	7.3359	83	2.27	028.32	0
028	Manganese (mg / kg (ppm))	0294	354.17	1.5400	292.33	21.016	7.3359	83	2.94	028.42	0
028	Manganese (mg / kg (ppm))	0169	371.00	34.000	292.33	21.016	7.3359	83	3.74	028.43	0
028	Manganese (mg / kg (ppm))	2178	273.50	167.00	292.33	21.016	7.3359	83	-0.90	028.33	1
028	Manganese (mg / kg (ppm))	0405	548.90	2.2000	292.33	21.016	7.3359	83	12.21	028.41	2
031	Phosphorus (%)	2066	0.67000	0.06000	0.86688	0.04307	0.01935	120	-4.57	031.53	0
031	Phosphorus (%)	0889	0.67880	0.00640	0.86688	0.04307	0.01935	120	-4.37	031.99	0
031	Phosphorus (%)	2192	0.76000	0.08000	0.86688	0.04307	0.01935	120	-2.48	031.01	0
031	Phosphorus (%)	0300	0.76200	0.00400	0.86688	0.04307	0.01935	120	-2.44	031.99	0
031	Phosphorus (%)	0885	0.76375	0.02210	0.86688	0.04307	0.01935	120	-2.39	031.01	0
031	Phosphorus (%)	0884	0.76500	0.01000	0.86688	0.04307	0.01935	120	-2.37	031.01	0
031	Phosphorus (%)	0685	0.78500	0.01000	0.86688	0.04307	0.01935	120	-1.90	031.41	0
031	Phosphorus (%)	0043	0.78500	0.03000	0.86688	0.04307	0.01935	120	-1.90	031.43	0
031	Phosphorus (%)	0154	0.79300	0.00300	0.86688	0.04307	0.01935	120	-1.72	031.52	0
031	Phosphorus (%)	0278	0.80000	0.06000	0.86688	0.04307	0.01935	120	-1.55	031.42	0
031	Phosphorus (%)	0870	0.80765	0.04310	0.86688	0.04307	0.01935	120	-1.38	031.42	0
031	Phosphorus (%)	0536	0.81000	0.00000	0.86688	0.04307	0.01935	120	-1.32	031.06	0
031	Phosphorus (%)	0910	0.81000	0.02000	0.86688	0.04307	0.01935	120	-1.32	031.41	0
031	Phosphorus (%)	0874	0.81285	0.00330	0.86688	0.04307	0.01935	120	-1.25	031.01	0
031	Phosphorus (%)	0108	0.82000	0.00000	0.86688	0.04307	0.01935	120	-1.09	031.01	0
031	Phosphorus (%)	0511	0.82000	0.08000	0.86688	0.04307	0.01935	120	-1.09	031.01	0
031	Phosphorus (%)	0019	0.82000	0.04000	0.86688	0.04307	0.01935	120	-1.09	031.41	0
031	Phosphorus (%)	2089	0.82000	0.04000	0.86688	0.04307	0.01935	120	-1.09	031.43	0
031	Phosphorus (%)	0045	0.82300	0.00400	0.86688	0.04307	0.01935	120	-1.02	031.42	0
031	Phosphorus (%)	0010	0.82500	0.01000	0.86688	0.04307	0.01935	120	-0.97	031.43	0
031	Phosphorus (%)	0918	0.82950	0.07500	0.86688	0.04307	0.01935	120	-0.87	031.43	0
031	Phosphorus (%)	0675	0.83000	0.02000	0.86688	0.04307	0.01935	120	-0.86	031.43	0
031	Phosphorus (%)	0100	0.83000	0.02000	0.86688	0.04307	0.01935	120	-0.86	031.99	0
031	Phosphorus (%)	0560	0.83035	0.00270	0.86688	0.04307	0.01935	120	-0.85	031.42	0
031	Phosphorus (%)	0345	0.83300	0.01000	0.86688	0.04307	0.01935	120	-0.79	031.43	0
031	Phosphorus (%)	0037	0.83500	0.01000	0.86688	0.04307	0.01935	120	-0.74	031.42	0
031	Phosphorus (%)	0035	0.83500	0.03000	0.86688	0.04307	0.01935	120	-0.74	031.43	0
031	Phosphorus (%)	0186	0.83750	0.02500	0.86688	0.04307	0.01935	120	-0.68	031.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	0187	0.83885	0.00190	0.86688	0.04307	0.01935	120	-0.65	031.42	0
031	Phosphorus (%)	0042	0.83950	0.05700	0.86688	0.04307	0.01935	120	-0.64	031.42	0
031	Phosphorus (%)	0897	0.84000	0.00000	0.86688	0.04307	0.01935	120	-0.62	031.01	0
031	Phosphorus (%)	0297	0.84000	0.04000	0.86688	0.04307	0.01935	120	-0.62	031.43	0
031	Phosphorus (%)	0629	0.84000	0.02000	0.86688	0.04307	0.01935	120	-0.62	031.43	0
031	Phosphorus (%)	0034	0.84155	0.02110	0.86688	0.04307	0.01935	120	-0.59	031.43	0
031	Phosphorus (%)	0035	0.84500	0.01000	0.86688	0.04307	0.01935	120	-0.51	031.01	0
031	Phosphorus (%)	0233	0.84500	0.01000	0.86688	0.04307	0.01935	120	-0.51	031.01	0
031	Phosphorus (%)	0144	0.84500	0.05000	0.86688	0.04307	0.01935	120	-0.51	031.41	0
031	Phosphorus (%)	0520	0.84500	0.03000	0.86688	0.04307	0.01935	120	-0.51	031.41	0
031	Phosphorus (%)	0848	0.84500	0.01000	0.86688	0.04307	0.01935	120	-0.51	031.41	0
031	Phosphorus (%)	2109	0.84800	0.01800	0.86688	0.04307	0.01935	120	-0.44	031.01	0
031	Phosphorus (%)	0590	0.85000	0.00000	0.86688	0.04307	0.01935	120	-0.39	031.01	0
031	Phosphorus (%)	0868	0.85000	0.02000	0.86688	0.04307	0.01935	120	-0.39	031.01	0
031	Phosphorus (%)	0294	0.85000	0.02000	0.86688	0.04307	0.01935	120	-0.39	031.42	0
031	Phosphorus (%)	0353	0.85000	0.04000	0.86688	0.04307	0.01935	120	-0.39	031.43	0
031	Phosphorus (%)	0229	0.85050	0.04300	0.86688	0.04307	0.01935	120	-0.38	031.41	0
031	Phosphorus (%)	0186	0.85200	0.00800	0.86688	0.04307	0.01935	120	-0.35	031.52	0
031	Phosphorus (%)	2126	0.85500	0.01000	0.86688	0.04307	0.01935	120	-0.28	031.01	0
031	Phosphorus (%)	0968	0.85850	0.00300	0.86688	0.04307	0.01935	120	-0.19	031.43	0
031	Phosphorus (%)	0098	0.85885	0.00570	0.86688	0.04307	0.01935	120	-0.19	031.41	0
031	Phosphorus (%)	0693	0.85950	0.04900	0.86688	0.04307	0.01935	120	-0.17	031.42	0
031	Phosphorus (%)	0689	0.86000	0.04000	0.86688	0.04307	0.01935	120	-0.16	031.01	0
031	Phosphorus (%)	2181	0.86000	0.02000	0.86688	0.04307	0.01935	120	-0.16	031.01	0
031	Phosphorus (%)	0955	0.86100	0.00400	0.86688	0.04307	0.01935	120	-0.14	031.44	0
031	Phosphorus (%)	0006	0.86400	0.02600	0.86688	0.04307	0.01935	120	-0.07	031.43	0
031	Phosphorus (%)	0609	0.86500	0.01000	0.86688	0.04307	0.01935	120	-0.04	031.01	0
031	Phosphorus (%)	0878	0.86500	0.01000	0.86688	0.04307	0.01935	120	-0.04	031.01	0
031	Phosphorus (%)	0939	0.86500	0.01000	0.86688	0.04307	0.01935	120	-0.04	031.01	0
031	Phosphorus (%)	0049	0.86500	0.07000	0.86688	0.04307	0.01935	120	-0.04	031.41	0
031	Phosphorus (%)	2145	0.86500	0.01000	0.86688	0.04307	0.01935	120	-0.04	031.41	0
031	Phosphorus (%)	0035	0.86500	0.01000	0.86688	0.04307	0.01935	120	-0.04	031.42	0
031	Phosphorus (%)	0692	0.86500	0.03000	0.86688	0.04307	0.01935	120	-0.04	031.42	0
031	Phosphorus (%)	0027	0.86500	0.03000	0.86688	0.04307	0.01935	120	-0.04	031.43	0
031	Phosphorus (%)	0142	0.86645	0.00330	0.86688	0.04307	0.01935	120	-0.01	031.01	0
031	Phosphorus (%)	0849	0.86900	0.00200	0.86688	0.04307	0.01935	120	0.05	031.01	0
031	Phosphorus (%)	0354	0.87000	0.00000	0.86688	0.04307	0.01935	120	0.07	031.01	0
031	Phosphorus (%)	0510	0.87000	0.04000	0.86688	0.04307	0.01935	120	0.07	031.43	0
031	Phosphorus (%)	0242	0.87000	0.00000	0.86688	0.04307	0.01935	120	0.07	031.99	0
031	Phosphorus (%)	0683	0.87500	0.01000	0.86688	0.04307	0.01935	120	0.19	031.01	0
031	Phosphorus (%)	0202	0.87550	0.00300	0.86688	0.04307	0.01935	120	0.20	031.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	0001	0.87600	0.01000	0.86688	0.04307	0.01935	120	0.21	031.03	0
031	Phosphorus (%)	2009	0.87655	0.00170	0.86688	0.04307	0.01935	120	0.22	031.01	0
031	Phosphorus (%)	0357	0.87820	0.02840	0.86688	0.04307	0.01935	120	0.26	031.42	0
031	Phosphorus (%)	2195	0.88000	0.00000	0.86688	0.04307	0.01935	120	0.30	031.01	0
031	Phosphorus (%)	0047	0.88000	0.06000	0.86688	0.04307	0.01935	120	0.30	031.42	0
031	Phosphorus (%)	0682	0.88000	0.00000	0.86688	0.04307	0.01935	120	0.30	031.43	0
031	Phosphorus (%)	0726	0.88080	0.00220	0.86688	0.04307	0.01935	120	0.32	031.42	0
031	Phosphorus (%)	0508	0.88240	0.03380	0.86688	0.04307	0.01935	120	0.36	031.43	0
031	Phosphorus (%)	0036	0.88410	0.01080	0.86688	0.04307	0.01935	120	0.40	031.03	0
031	Phosphorus (%)	0653	0.88450	0.00700	0.86688	0.04307	0.01935	120	0.41	031.01	0
031	Phosphorus (%)	0723	0.88490	0.00000	0.86688	0.04307	0.01935	120	0.42	031.01	0
031	Phosphorus (%)	0629	0.88500	0.01000	0.86688	0.04307	0.01935	120	0.42	031.01	0
031	Phosphorus (%)	0413	0.88500	0.01000	0.86688	0.04307	0.01935	120	0.42	031.41	0
031	Phosphorus (%)	0016	0.88550	0.06300	0.86688	0.04307	0.01935	120	0.43	031.01	0
031	Phosphorus (%)	0043	0.88800	0.00600	0.86688	0.04307	0.01935	120	0.49	031.02	0
031	Phosphorus (%)	0881	0.88950	0.00100	0.86688	0.04307	0.01935	120	0.53	031.01	0
031	Phosphorus (%)	0139	0.89000	0.01400	0.86688	0.04307	0.01935	120	0.54	031.41	0
031	Phosphorus (%)	0011	0.89310	0.03340	0.86688	0.04307	0.01935	120	0.61	031.41	0
031	Phosphorus (%)	0036	0.89385	0.01470	0.86688	0.04307	0.01935	120	0.63	031.43	0
031	Phosphorus (%)	0948	0.89500	0.00400	0.86688	0.04307	0.01935	120	0.65	031.01	0
031	Phosphorus (%)	2188	0.89500	0.01000	0.86688	0.04307	0.01935	120	0.65	031.01	0
031	Phosphorus (%)	2053	0.89500	0.01000	0.86688	0.04307	0.01935	120	0.65	031.02	0
031	Phosphorus (%)	0169	0.89500	0.05000	0.86688	0.04307	0.01935	120	0.65	031.03	0
031	Phosphorus (%)	0263	0.89635	0.00210	0.86688	0.04307	0.01935	120	0.68	031.42	0
031	Phosphorus (%)	0407	0.89985	0.00150	0.86688	0.04307	0.01935	120	0.77	031.41	0
031	Phosphorus (%)	2196	0.90000	0.00000	0.86688	0.04307	0.01935	120	0.77	031.01	0
031	Phosphorus (%)	0171	0.90000	0.00000	0.86688	0.04307	0.01935	120	0.77	031.41	0
031	Phosphorus (%)	0208	0.90100	0.03600	0.86688	0.04307	0.01935	120	0.79	031.01	0
031	Phosphorus (%)	0190	0.90500	0.01000	0.86688	0.04307	0.01935	120	0.89	031.42	0
031	Phosphorus (%)	0160	0.90555	0.00550	0.86688	0.04307	0.01935	120	0.90	031.44	0
031	Phosphorus (%)	0014	0.90600	0.00000	0.86688	0.04307	0.01935	120	0.91	031.42	0
031	Phosphorus (%)	0032	0.90675	0.03310	0.86688	0.04307	0.01935	120	0.93	031.43	0
031	Phosphorus (%)	0175	0.91000	0.02000	0.86688	0.04307	0.01935	120	1.00	031.01	0
031	Phosphorus (%)	0505	0.91000	0.02000	0.86688	0.04307	0.01935	120	1.00	031.02	0
031	Phosphorus (%)	0205	0.91000	0.01800	0.86688	0.04307	0.01935	120	1.00	031.42	0
031	Phosphorus (%)	0098	0.91050	0.00940	0.86688	0.04307	0.01935	120	1.01	031.43	0
031	Phosphorus (%)	0572	0.91100	0.00400	0.86688	0.04307	0.01935	120	1.02	031.53	0
031	Phosphorus (%)	0895	0.91500	0.01000	0.86688	0.04307	0.01935	120	1.12	031.00	0
031	Phosphorus (%)	0512	0.92605	0.03970	0.86688	0.04307	0.01935	120	1.37	031.43	0
031	Phosphorus (%)	2006	0.92900	0.01200	0.86688	0.04307	0.01935	120	1.44	031.01	0
031	Phosphorus (%)	0208	0.93950	0.07900	0.86688	0.04307	0.01935	120	1.69	031.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	0083	0.95000	0.02000	0.86688	0.04307	0.01935	120	1.93	031.43	0
031	Phosphorus (%)	0358	0.96000	0.00000	0.86688	0.04307	0.01935	120	2.16	031.41	0
031	Phosphorus (%)	0598	0.96150	0.01760	0.86688	0.04307	0.01935	120	2.20	031.41	0
031	Phosphorus (%)	2022	0.96500	0.01000	0.86688	0.04307	0.01935	120	2.28	031.01	0
031	Phosphorus (%)	0226	0.98000	0.00000	0.86688	0.04307	0.01935	120	2.63	031.41	0
031	Phosphorus (%)	0089	0.98000	0.00000	0.86688	0.04307	0.01935	120	2.63	031.43	0
031	Phosphorus (%)	0425	1.0000	0.02000	0.86688	0.04307	0.01935	120	3.09	031.43	0
031	Phosphorus (%)	0405	1.0350	0.03000	0.86688	0.04307	0.01935	120	3.90	031.41	0
031	Phosphorus (%)	0123	1.0700	0.00000	0.86688	0.04307	0.01935	120	4.72	031.41	0
031	Phosphorus (%)	0003	1.1250	0.03000	0.86688	0.04307	0.01935	120	5.99	031.41	0
031	Phosphorus (%)	0152	0.75000	0.10000	0.86688	0.04307	0.01935	120	-2.71	031.01	1
031	Phosphorus (%)	0687	0.95000	0.10000	0.86688	0.04307	0.01935	120	1.93	031.01	1
032	Potassium (%)	0868	0.18500	0.01000	1.2763	0.06359	0.03951	89	-17.16	032.31	0
032	Potassium (%)	0889	0.93475	0.01570	1.2763	0.06359	0.03951	89	-5.37	032.99	0
032	Potassium (%)	2066	1.0050	0.07000	1.2763	0.06359	0.03951	89	-4.27	032.53	0
032	Potassium (%)	0609	1.0450	0.03000	1.2763	0.06359	0.03951	89	-3.64	032.32	0
032	Potassium (%)	0536	1.0900	0.06000	1.2763	0.06359	0.03951	89	-2.93	032.31	0
032	Potassium (%)	0003	1.1600	0.02000	1.2763	0.06359	0.03951	89	-1.83	032.41	0
032	Potassium (%)	0511	1.1850	0.11000	1.2763	0.06359	0.03951	89	-1.44	032.41	0
032	Potassium (%)	0019	1.1900	0.06000	1.2763	0.06359	0.03951	89	-1.36	032.41	0
032	Potassium (%)	0035	1.1900	0.04000	1.2763	0.06359	0.03951	89	-1.36	032.43	0
032	Potassium (%)	0278	1.2000	0.02000	1.2763	0.06359	0.03951	89	-1.20	032.42	0
032	Potassium (%)	0014	1.2050	0.03000	1.2763	0.06359	0.03951	89	-1.12	032.31	0
032	Potassium (%)	0175	1.2100	0.00000	1.2763	0.06359	0.03951	89	-1.04	032.31	0
032	Potassium (%)	0425	1.2100	0.04000	1.2763	0.06359	0.03951	89	-1.04	032.43	0
032	Potassium (%)	0693	1.2105	0.02100	1.2763	0.06359	0.03951	89	-1.03	032.42	0
032	Potassium (%)	2089	1.2150	0.01000	1.2763	0.06359	0.03951	89	-0.96	032.43	0
032	Potassium (%)	0354	1.2250	0.01000	1.2763	0.06359	0.03951	89	-0.81	032.31	0
032	Potassium (%)	0049	1.2250	0.07000	1.2763	0.06359	0.03951	89	-0.81	032.41	0
032	Potassium (%)	0870	1.2290	0.03000	1.2763	0.06359	0.03951	89	-0.74	032.42	0
032	Potassium (%)	0918	1.2295	0.04300	1.2763	0.06359	0.03951	89	-0.74	032.43	0
032	Potassium (%)	0169	1.2300	0.10000	1.2763	0.06359	0.03951	89	-0.73	032.32	0
032	Potassium (%)	0100	1.2350	0.03000	1.2763	0.06359	0.03951	89	-0.65	032.99	0
032	Potassium (%)	0171	1.2370	0.00400	1.2763	0.06359	0.03951	89	-0.62	032.41	0
032	Potassium (%)	0229	1.2385	0.04500	1.2763	0.06359	0.03951	89	-0.59	032.41	0
032	Potassium (%)	0629	1.2400	0.00000	1.2763	0.06359	0.03951	89	-0.57	032.31	0
032	Potassium (%)	0881	1.2400	0.01400	1.2763	0.06359	0.03951	89	-0.57	032.43	0
032	Potassium (%)	0034	1.2425	0.00500	1.2763	0.06359	0.03951	89	-0.53	032.43	0
032	Potassium (%)	0208	1.2430	0.05400	1.2763	0.06359	0.03951	89	-0.52	032.41	0
032	Potassium (%)	0154	1.2469	0.05180	1.2763	0.06359	0.03951	89	-0.46	032.52	0
032	Potassium (%)	0098	1.2500	0.00000	1.2763	0.06359	0.03951	89	-0.41	032.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
032	Potassium (%)	0510	1.2500	0.02000	1.2763	0.06359	0.03951	89	-0.41	032.43	0
032	Potassium (%)	0011	1.2538	0.04720	1.2763	0.06359	0.03951	89	-0.35	032.41	0
032	Potassium (%)	0560	1.2541	0.00530	1.2763	0.06359	0.03951	89	-0.35	032.42	0
032	Potassium (%)	2188	1.2550	0.05000	1.2763	0.06359	0.03951	89	-0.33	032.08	0
032	Potassium (%)	0144	1.2550	0.13000	1.2763	0.06359	0.03951	89	-0.33	032.41	0
032	Potassium (%)	0297	1.2550	0.07000	1.2763	0.06359	0.03951	89	-0.33	032.43	0
032	Potassium (%)	0353	1.2550	0.01000	1.2763	0.06359	0.03951	89	-0.33	032.43	0
032	Potassium (%)	0629	1.2550	0.09000	1.2763	0.06359	0.03951	89	-0.33	032.43	0
032	Potassium (%)	0874	1.2600	0.06000	1.2763	0.06359	0.03951	89	-0.26	032.31	0
032	Potassium (%)	0910	1.2600	0.02000	1.2763	0.06359	0.03951	89	-0.26	032.41	0
032	Potassium (%)	0037	1.2600	0.02000	1.2763	0.06359	0.03951	89	-0.26	032.42	0
032	Potassium (%)	0692	1.2600	0.16000	1.2763	0.06359	0.03951	89	-0.26	032.42	0
032	Potassium (%)	0955	1.2600	0.02000	1.2763	0.06359	0.03951	89	-0.26	032.44	0
032	Potassium (%)	0043	1.2650	0.01000	1.2763	0.06359	0.03951	89	-0.18	032.43	0
032	Potassium (%)	0042	1.2700	0.04000	1.2763	0.06359	0.03951	89	-0.10	032.42	0
032	Potassium (%)	0345	1.2700	0.02000	1.2763	0.06359	0.03951	89	-0.10	032.43	0
032	Potassium (%)	0045	1.2750	0.09000	1.2763	0.06359	0.03951	89	-0.02	032.42	0
032	Potassium (%)	0294	1.2750	0.07000	1.2763	0.06359	0.03951	89	-0.02	032.42	0
032	Potassium (%)	0512	1.2790	0.05200	1.2763	0.06359	0.03951	89	0.04	032.43	0
032	Potassium (%)	0187	1.2800	0.00000	1.2763	0.06359	0.03951	89	0.06	032.42	0
032	Potassium (%)	0242	1.2800	0.00000	1.2763	0.06359	0.03951	89	0.06	032.99	0
032	Potassium (%)	0036	1.2805	0.00330	1.2763	0.06359	0.03951	89	0.07	032.43	0
032	Potassium (%)	0405	1.2825	0.01500	1.2763	0.06359	0.03951	89	0.10	032.41	0
032	Potassium (%)	0205	1.2900	0.02000	1.2763	0.06359	0.03951	89	0.22	032.42	0
032	Potassium (%)	0001	1.2915	0.01900	1.2763	0.06359	0.03951	89	0.24	032.99	0
032	Potassium (%)	0186	1.2950	0.01000	1.2763	0.06359	0.03951	89	0.29	032.52	0
032	Potassium (%)	0948	1.2955	0.00500	1.2763	0.06359	0.03951	89	0.30	032.31	0
032	Potassium (%)	0208	1.3000	0.10000	1.2763	0.06359	0.03951	89	0.37	032.31	0
032	Potassium (%)	0878	1.3000	0.00000	1.2763	0.06359	0.03951	89	0.37	032.41	0
032	Potassium (%)	0010	1.3050	0.07000	1.2763	0.06359	0.03951	89	0.45	032.43	0
032	Potassium (%)	0142	1.3090	0.00500	1.2763	0.06359	0.03951	89	0.51	032.31	0
032	Potassium (%)	2053	1.3100	0.02000	1.2763	0.06359	0.03951	89	0.53	032.31	0
032	Potassium (%)	0682	1.3100	0.00000	1.2763	0.06359	0.03951	89	0.53	032.43	0
032	Potassium (%)	0508	1.3130	0.04830	1.2763	0.06359	0.03951	89	0.58	032.43	0
032	Potassium (%)	0139	1.3130	0.02000	1.2763	0.06359	0.03951	89	0.58	032.41	0
032	Potassium (%)	0027	1.3155	0.05700	1.2763	0.06359	0.03951	89	0.62	032.43	0
032	Potassium (%)	0160	1.3160	0.04330	1.2763	0.06359	0.03951	89	0.62	032.44	0
032	Potassium (%)	0358	1.3200	0.04000	1.2763	0.06359	0.03951	89	0.69	032.41	0
032	Potassium (%)	0413	1.3200	0.00000	1.2763	0.06359	0.03951	89	0.69	032.41	0
032	Potassium (%)	0598	1.3235	0.01900	1.2763	0.06359	0.03951	89	0.74	032.41	0
032	Potassium (%)	0939	1.3250	0.05000	1.2763	0.06359	0.03951	89	0.77	032.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
032	Potassium (%)	0035	1.3250	0.03000	1.2763	0.06359	0.03951	89	0.77	032.42	0
032	Potassium (%)	0098	1.3300	0.10000	1.2763	0.06359	0.03951	89	0.84	032.41	0
032	Potassium (%)	0520	1.3300	0.06000	1.2763	0.06359	0.03951	89	0.84	032.41	0
032	Potassium (%)	0186	1.3350	0.01000	1.2763	0.06359	0.03951	89	0.92	032.42	0
032	Potassium (%)	0407	1.3360	0.00090	1.2763	0.06359	0.03951	89	0.94	032.41	0
032	Potassium (%)	0357	1.3388	0.07520	1.2763	0.06359	0.03951	89	0.98	032.42	0
032	Potassium (%)	0675	1.3400	0.02000	1.2763	0.06359	0.03951	89	1.00	032.43	0
032	Potassium (%)	0726	1.3500	0.00400	1.2763	0.06359	0.03951	89	1.16	032.42	0
032	Potassium (%)	0047	1.3550	0.07000	1.2763	0.06359	0.03951	89	1.24	032.42	0
032	Potassium (%)	0202	1.3590	0.01800	1.2763	0.06359	0.03951	89	1.30	032.43	0
032	Potassium (%)	0612	1.3650	0.05000	1.2763	0.06359	0.03951	89	1.40	032.33	0
032	Potassium (%)	0190	1.3700	0.04000	1.2763	0.06359	0.03951	89	1.47	032.42	0
032	Potassium (%)	0263	1.3789	0.03570	1.2763	0.06359	0.03951	89	1.61	032.42	0
032	Potassium (%)	0572	1.3800	0.14000	1.2763	0.06359	0.03951	89	1.63	032.53	0
032	Potassium (%)	0083	1.3950	0.03000	1.2763	0.06359	0.03951	89	1.87	032.43	0
032	Potassium (%)	0968	1.4240	0.02000	1.2763	0.06359	0.03951	89	2.32	032.43	0
032	Potassium (%)	0226	1.4450	0.01000	1.2763	0.06359	0.03951	89	2.65	032.41	0
032	Potassium (%)	0089	1.4800	0.00000	1.2763	0.06359	0.03951	89	3.20	032.43	0
032	Potassium (%)	0032	11.860	0.28000	1.2763	0.06359	0.03951	89	166.45	032.43	0
032	Potassium (%)	2195	12.720	325.00	1.2763	0.06359	0.03951	89	20016.88	032.31	2
033	Salt as chloride (%)	0723	0.30680	0.00000	1.1671	0.08694	0.02176	47	-9.90	033.00	0
033	Salt as chloride (%)	0868	0.53500	0.00200	1.1671	0.08694	0.02176	47	-7.27	033.05	0
033	Salt as chloride (%)	0354	0.72500	0.01000	1.1671	0.08694	0.02176	47	-5.09	033.01	0
033	Salt as chloride (%)	2076	0.85000	0.08000	1.1671	0.08694	0.02176	47	-3.65	033.03	0
033	Salt as chloride (%)	2181	1.0100	0.06000	1.1671	0.08694	0.02176	47	-1.81	033.99	0
033	Salt as chloride (%)	0171	1.0200	0.06000	1.1671	0.08694	0.02176	47	-1.69	033.00	0
033	Salt as chloride (%)	0144	1.0450	0.09000	1.1671	0.08694	0.02176	47	-1.40	033.03	0
033	Salt as chloride (%)	0229	1.0580	0.00200	1.1671	0.08694	0.02176	47	-1.26	033.01	0
033	Salt as chloride (%)	0897	1.0650	0.03000	1.1671	0.08694	0.02176	47	-1.17	033.00	0
033	Salt as chloride (%)	0536	1.0650	0.05000	1.1671	0.08694	0.02176	47	-1.17	033.99	0
033	Salt as chloride (%)	0693	1.1040	0.11400	1.1671	0.08694	0.02176	47	-0.73	033.00	0
033	Salt as chloride (%)	0878	1.1200	0.02000	1.1671	0.08694	0.02176	47	-0.54	033.00	0
033	Salt as chloride (%)	0884	1.1200	0.02000	1.1671	0.08694	0.02176	47	-0.54	033.00	0
033	Salt as chloride (%)	2073	1.1500	0.02000	1.1671	0.08694	0.02176	47	-0.20	033.99	0
033	Salt as chloride (%)	0309	1.1581	0.02470	1.1671	0.08694	0.02176	47	-0.10	033.00	0
033	Salt as chloride (%)	0689	1.1600	0.02000	1.1671	0.08694	0.02176	47	-0.08	033.05	0
033	Salt as chloride (%)	0910	1.1600	0.02000	1.1671	0.08694	0.02176	47	-0.08	033.99	0
033	Salt as chloride (%)	0675	1.1700	0.02000	1.1671	0.08694	0.02176	47	0.03	033.00	0
033	Salt as chloride (%)	0004	1.1700	0.02000	1.1671	0.08694	0.02176	47	0.03	033.01	0
033	Salt as chloride (%)	0425	1.1700	0.00000	1.1671	0.08694	0.02176	47	0.03	033.01	0
033	Salt as chloride (%)	2109	1.1700	0.03200	1.1671	0.08694	0.02176	47	0.03	033.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
033	Salt as chloride (%)	0353	1.1800	0.02000	1.1671	0.08694	0.02176	47	0.15	033.00	0
033	Salt as chloride (%)	0226	1.1800	0.02000	1.1671	0.08694	0.02176	47	0.15	033.01	0
033	Salt as chloride (%)	0590	1.1800	0.00000	1.1671	0.08694	0.02176	47	0.15	033.01	0
033	Salt as chloride (%)	0683	1.1800	0.00000	1.1671	0.08694	0.02176	47	0.15	033.01	0
033	Salt as chloride (%)	2009	1.1803	0.00280	1.1671	0.08694	0.02176	47	0.15	033.00	0
033	Salt as chloride (%)	0242	1.1850	0.03000	1.1671	0.08694	0.02176	47	0.21	033.01	0
033	Salt as chloride (%)	2146	1.1850	0.01000	1.1671	0.08694	0.02176	47	0.21	033.01	0
033	Salt as chloride (%)	0100	1.1950	0.01000	1.1671	0.08694	0.02176	47	0.32	033.01	0
033	Salt as chloride (%)	0413	1.1950	0.01000	1.1671	0.08694	0.02176	47	0.32	033.01	0
033	Salt as chloride (%)	0948	1.2020	0.00200	1.1671	0.08694	0.02176	47	0.40	033.01	0
033	Salt as chloride (%)	0885	1.2078	0.01930	1.1671	0.08694	0.02176	47	0.47	033.00	0
033	Salt as chloride (%)	0644	1.2097	0.03780	1.1671	0.08694	0.02176	47	0.49	033.01	0
033	Salt as chloride (%)	0205	1.2100	0.00800	1.1671	0.08694	0.02176	47	0.49	033.01	0
033	Salt as chloride (%)	0407	1.2135	0.00300	1.1671	0.08694	0.02176	47	0.53	033.01	0
033	Salt as chloride (%)	0098	1.2150	0.01000	1.1671	0.08694	0.02176	47	0.55	033.01	0
033	Salt as chloride (%)	0175	1.2250	0.01000	1.1671	0.08694	0.02176	47	0.67	033.01	0
033	Salt as chloride (%)	0278	1.2300	0.02000	1.1671	0.08694	0.02176	47	0.72	033.01	0
033	Salt as chloride (%)	0510	1.2300	0.02000	1.1671	0.08694	0.02176	47	0.72	033.01	0
033	Salt as chloride (%)	0874	1.2350	0.03000	1.1671	0.08694	0.02176	47	0.78	033.01	0
033	Salt as chloride (%)	0539	1.2400	0.00000	1.1671	0.08694	0.02176	47	0.84	033.00	0
033	Salt as chloride (%)	0629	1.2500	0.02000	1.1671	0.08694	0.02176	47	0.95	033.01	0
033	Salt as chloride (%)	0045	1.2600	0.00000	1.1671	0.08694	0.02176	47	1.07	033.00	0
033	Salt as chloride (%)	0881	1.2765	0.01500	1.1671	0.08694	0.02176	47	1.26	033.00	0
033	Salt as chloride (%)	0682	1.2900	0.00000	1.1671	0.08694	0.02176	47	1.41	033.99	0
033	Salt as chloride (%)	0358	1.9850	0.03000	1.1671	0.08694	0.02176	47	9.41	033.99	0
033	Salt as chloride (%)	0638	1.9900	0.00000	1.1671	0.08694	0.02176	47	9.46	033.01	0
033	Salt as chloride (%)	2043	1.7150	0.15000	1.1671	0.08694	0.02176	47	6.30	033.05	1
034	Selenium (mg / kg (ppm))	2195	0.68200	0.05800	2.0809	0.56066	0.11236	21	-2.50	034.31	0
034	Selenium (mg / kg (ppm))	0045	1.1950	0.29000	2.0809	0.56066	0.11236	21	-1.58	034.04	0
034	Selenium (mg / kg (ppm))	0171	1.4600	0.06000	2.0809	0.56066	0.11236	21	-1.11	034.04	0
034	Selenium (mg / kg (ppm))	0675	1.6150	0.15000	2.0809	0.56066	0.11236	21	-0.83	034.43	0
034	Selenium (mg / kg (ppm))	2066	1.7400	0.04000	2.0809	0.56066	0.11236	21	-0.61	034.53	0
034	Selenium (mg / kg (ppm))	0169	1.7500	0.04000	2.0809	0.56066	0.11236	21	-0.59	034.04	0
034	Selenium (mg / kg (ppm))	0955	1.7800	0.38000	2.0809	0.56066	0.11236	21	-0.54	034.43	0
034	Selenium (mg / kg (ppm))	0512	1.8755	0.03700	2.0809	0.56066	0.11236	21	-0.37	034.99	0
034	Selenium (mg / kg (ppm))	0939	1.9350	0.07000	2.0809	0.56066	0.11236	21	-0.26	034.04	0
034	Selenium (mg / kg (ppm))	0047	1.9800	0.04000	2.0809	0.56066	0.11236	21	-0.18	034.52	0
034	Selenium (mg / kg (ppm))	0208	2.1850	0.03000	2.0809	0.56066	0.11236	21	0.19	034.52	0
034	Selenium (mg / kg (ppm))	0098	2.1990	0.22800	2.0809	0.56066	0.11236	21	0.21	034.53	0
034	Selenium (mg / kg (ppm))	0560	2.2340	0.06600	2.0809	0.56066	0.11236	21	0.27	034.53	0
034	Selenium (mg / kg (ppm))	0190	2.2350	0.13000	2.0809	0.56066	0.11236	21	0.27	034.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
034	Selenium (mg / kg (ppm))	0186	2.2600	0.02000	2.0809	0.56066	0.11236	21	0.32	034.52	0
034	Selenium (mg / kg (ppm))	0910	2.3150	0.35000	2.0809	0.56066	0.11236	21	0.42	034.52	0
034	Selenium (mg / kg (ppm))	0918	2.3338	0.05060	2.0809	0.56066	0.11236	21	0.45	034.53	0
034	Selenium (mg / kg (ppm))	0032	2.7850	0.03000	2.0809	0.56066	0.11236	21	1.26	034.53	0
034	Selenium (mg / kg (ppm))	0036	2.8000	0.20000	2.0809	0.56066	0.11236	21	1.28	034.43	0
034	Selenium (mg / kg (ppm))	0016	2.8150	0.09000	2.0809	0.56066	0.11236	21	1.31	034.52	0
034	Selenium (mg / kg (ppm))	0692	3.0000	0.00000	2.0809	0.56066	0.11236	21	1.64	034.42	0
035	Sodium (%)	0889	0.36045	0.01850	0.61983	0.04060	0.01577	86	-6.39	035.99	0
035	Sodium (%)	2192	0.51000	0.02000	0.61983	0.04060	0.01577	86	-2.71	035.43	0
035	Sodium (%)	0881	0.53200	0.00400	0.61983	0.04060	0.01577	86	-2.16	035.43	0
035	Sodium (%)	2188	0.53500	0.01000	0.61983	0.04060	0.01577	86	-2.09	035.01	0
035	Sodium (%)	0043	0.55450	0.00700	0.61983	0.04060	0.01577	86	-1.61	035.43	0
035	Sodium (%)	0175	0.55500	0.01000	0.61983	0.04060	0.01577	86	-1.60	035.31	0
035	Sodium (%)	0233	0.55500	0.01000	0.61983	0.04060	0.01577	86	-1.60	035.31	0
035	Sodium (%)	0035	0.56500	0.01000	0.61983	0.04060	0.01577	86	-1.35	035.43	0
035	Sodium (%)	0536	0.57000	0.00000	0.61983	0.04060	0.01577	86	-1.23	035.31	0
035	Sodium (%)	0510	0.57350	0.00500	0.61983	0.04060	0.01577	86	-1.14	035.43	0
035	Sodium (%)	0098	0.57500	0.01000	0.61983	0.04060	0.01577	86	-1.10	035.43	0
035	Sodium (%)	0353	0.57500	0.01000	0.61983	0.04060	0.01577	86	-1.10	035.43	0
035	Sodium (%)	0300	0.58000	0.02000	0.61983	0.04060	0.01577	86	-0.98	035.33	0
035	Sodium (%)	0278	0.58000	0.02000	0.61983	0.04060	0.01577	86	-0.98	035.42	0
035	Sodium (%)	0032	0.58000	0.01240	0.61983	0.04060	0.01577	86	-0.98	035.43	0
035	Sodium (%)	0345	0.58300	0.00400	0.61983	0.04060	0.01577	86	-0.91	035.43	0
035	Sodium (%)	0045	0.58650	0.01100	0.61983	0.04060	0.01577	86	-0.82	035.42	0
035	Sodium (%)	0689	0.59000	0.00000	0.61983	0.04060	0.01577	86	-0.73	035.31	0
035	Sodium (%)	2188	0.59000	0.02000	0.61983	0.04060	0.01577	86	-0.73	035.31	0
035	Sodium (%)	0144	0.59000	0.04000	0.61983	0.04060	0.01577	86	-0.73	035.41	0
035	Sodium (%)	0037	0.59500	0.01000	0.61983	0.04060	0.01577	86	-0.61	035.05	0
035	Sodium (%)	0019	0.59500	0.01000	0.61983	0.04060	0.01577	86	-0.61	035.41	0
035	Sodium (%)	0171	0.59500	0.01000	0.61983	0.04060	0.01577	86	-0.61	035.41	0
035	Sodium (%)	0154	0.59515	0.00550	0.61983	0.04060	0.01577	86	-0.61	035.52	0
035	Sodium (%)	0049	0.60000	0.06000	0.61983	0.04060	0.01577	86	-0.49	035.41	0
035	Sodium (%)	0297	0.60000	0.00000	0.61983	0.04060	0.01577	86	-0.49	035.43	0
035	Sodium (%)	0692	0.60000	0.10000	0.61983	0.04060	0.01577	86	-0.49	035.42	0
035	Sodium (%)	0560	0.60035	0.00590	0.61983	0.04060	0.01577	86	-0.48	035.42	0
035	Sodium (%)	0229	0.60450	0.00700	0.61983	0.04060	0.01577	86	-0.38	035.41	0
035	Sodium (%)	0629	0.60500	0.01000	0.61983	0.04060	0.01577	86	-0.37	035.31	0
035	Sodium (%)	0878	0.60500	0.03000	0.61983	0.04060	0.01577	86	-0.37	035.41	0
035	Sodium (%)	0910	0.60500	0.05000	0.61983	0.04060	0.01577	86	-0.37	035.41	0
035	Sodium (%)	0693	0.60850	0.00700	0.61983	0.04060	0.01577	86	-0.28	035.42	0
035	Sodium (%)	0294	0.61000	0.02000	0.61983	0.04060	0.01577	86	-0.24	035.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
035	Sodium (%)	0870	0.61000	0.02500	0.61983	0.04060	0.01577	86	-0.24	035.42	0
035	Sodium (%)	0205	0.61200	0.00400	0.61983	0.04060	0.01577	86	-0.19	035.42	0
035	Sodium (%)	0042	0.61250	0.05700	0.61983	0.04060	0.01577	86	-0.18	035.42	0
035	Sodium (%)	0006	0.61300	0.01400	0.61983	0.04060	0.01577	86	-0.17	035.43	0
035	Sodium (%)	0263	0.61425	0.00290	0.61983	0.04060	0.01577	86	-0.14	035.42	0
035	Sodium (%)	0520	0.61605	0.00490	0.61983	0.04060	0.01577	86	-0.09	035.41	0
035	Sodium (%)	0723	0.61630	0.00000	0.61983	0.04060	0.01577	86	-0.09	035.31	0
035	Sodium (%)	0208	0.61800	0.00200	0.61983	0.04060	0.01577	86	-0.05	035.41	0
035	Sodium (%)	0142	0.61960	0.00000	0.61983	0.04060	0.01577	86	-0.01	035.31	0
035	Sodium (%)	0868	0.62000	0.00000	0.61983	0.04060	0.01577	86	0.00	035.31	0
035	Sodium (%)	0169	0.62000	0.00000	0.61983	0.04060	0.01577	86	0.00	035.32	0
035	Sodium (%)	0035	0.62000	0.00000	0.61983	0.04060	0.01577	86	0.00	035.42	0
035	Sodium (%)	0682	0.62000	0.00000	0.61983	0.04060	0.01577	86	0.00	035.43	0
035	Sodium (%)	0918	0.62200	0.00200	0.61983	0.04060	0.01577	86	0.05	035.43	0
035	Sodium (%)	0598	0.62300	0.00780	0.61983	0.04060	0.01577	86	0.08	035.41	0
035	Sodium (%)	0413	0.62500	0.05000	0.61983	0.04060	0.01577	86	0.13	035.41	0
035	Sodium (%)	0208	0.62750	0.00300	0.61983	0.04060	0.01577	86	0.19	035.31	0
035	Sodium (%)	0187	0.62815	0.00030	0.61983	0.04060	0.01577	86	0.20	035.42	0
035	Sodium (%)	0098	0.63000	0.04000	0.61983	0.04060	0.01577	86	0.25	035.41	0
035	Sodium (%)	0100	0.63000	0.02000	0.61983	0.04060	0.01577	86	0.25	035.99	0
035	Sodium (%)	0242	0.63000	0.00000	0.61983	0.04060	0.01577	86	0.25	035.99	0
035	Sodium (%)	0108	0.63500	0.03000	0.61983	0.04060	0.01577	86	0.37	035.05	0
035	Sodium (%)	0609	0.63500	0.01000	0.61983	0.04060	0.01577	86	0.37	035.32	0
035	Sodium (%)	0425	0.63500	0.01000	0.61983	0.04060	0.01577	86	0.37	035.43	0
035	Sodium (%)	0948	0.63550	0.00300	0.61983	0.04060	0.01577	86	0.39	035.31	0
035	Sodium (%)	0407	0.63855	0.00150	0.61983	0.04060	0.01577	86	0.46	035.41	0
035	Sodium (%)	0939	0.64000	0.02000	0.61983	0.04060	0.01577	86	0.50	035.31	0
035	Sodium (%)	2089	0.64000	0.02000	0.61983	0.04060	0.01577	86	0.50	035.43	0
035	Sodium (%)	0508	0.64180	0.05000	0.61983	0.04060	0.01577	86	0.54	035.43	0
035	Sodium (%)	2196	0.64300	0.00000	0.61983	0.04060	0.01577	86	0.57	035.31	0
035	Sodium (%)	2006	0.64450	0.00100	0.61983	0.04060	0.01577	86	0.61	035.01	0
035	Sodium (%)	0572	0.64450	0.00300	0.61983	0.04060	0.01577	86	0.61	035.53	0
035	Sodium (%)	0675	0.64500	0.03000	0.61983	0.04060	0.01577	86	0.62	035.43	0
035	Sodium (%)	0186	0.64700	0.02200	0.61983	0.04060	0.01577	86	0.67	035.42	0
035	Sodium (%)	0874	0.65200	0.07800	0.61983	0.04060	0.01577	86	0.79	035.31	0
035	Sodium (%)	0955	0.65400	0.06000	0.61983	0.04060	0.01577	86	0.84	035.41	0
035	Sodium (%)	0590	0.65500	0.03000	0.61983	0.04060	0.01577	86	0.87	035.99	0
035	Sodium (%)	0202	0.65850	0.01300	0.61983	0.04060	0.01577	86	0.95	035.43	0
035	Sodium (%)	0139	0.66200	0.01000	0.61983	0.04060	0.01577	86	1.04	035.41	0
035	Sodium (%)	0629	0.66500	0.05000	0.61983	0.04060	0.01577	86	1.11	035.43	0
035	Sodium (%)	0405	0.66950	0.01900	0.61983	0.04060	0.01577	86	1.22	035.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
035	Sodium (%)	0354	0.67000	0.00000	0.61983	0.04060	0.01577	86	1.24	035.31	0
035	Sodium (%)	0186	0.67250	0.01500	0.61983	0.04060	0.01577	86	1.30	035.52	0
035	Sodium (%)	0190	0.68000	0.02000	0.61983	0.04060	0.01577	86	1.48	035.42	0
035	Sodium (%)	0726	0.69470	0.00460	0.61983	0.04060	0.01577	86	1.84	035.42	0
035	Sodium (%)	2126	0.69500	0.01000	0.61983	0.04060	0.01577	86	1.85	035.05	0
035	Sodium (%)	0083	0.69500	0.01000	0.61983	0.04060	0.01577	86	1.85	035.43	0
035	Sodium (%)	0089	0.70000	0.00000	0.61983	0.04060	0.01577	86	1.97	035.43	0
035	Sodium (%)	0226	0.73600	0.00600	0.61983	0.04060	0.01577	86	2.86	035.41	0
035	Sodium (%)	0358	0.78500	0.01000	0.61983	0.04060	0.01577	86	4.07	035.41	0
035	Sodium (%)	0123	1.0100	0.02000	0.61983	0.04060	0.01577	86	9.61	035.41	0
035	Sodium (%)	0868	1.1800	0.00000	0.61983	0.04060	0.01577	86	13.80	035.01	0
035	Sodium (%)	2195	6,202.5	53.000	0.61983	0.04060	0.01577	86	152750.60	035.31	2
036	Sulfur (%)	0910	0.25245	0.00010	0.31122	0.02175	0.01107	34	-2.70	036.52	0
036	Sulfur (%)	0870	0.26775	0.01590	0.31122	0.02175	0.01107	34	-2.00	036.42	0
036	Sulfur (%)	0353	0.28000	0.02000	0.31122	0.02175	0.01107	34	-1.44	036.43	0
036	Sulfur (%)	0035	0.28500	0.03000	0.31122	0.02175	0.01107	34	-1.21	036.43	0
036	Sulfur (%)	0045	0.29200	0.01800	0.31122	0.02175	0.01107	34	-0.88	036.42	0
036	Sulfur (%)	0708	0.29250	0.00100	0.31122	0.02175	0.01107	34	-0.86	036.42	0
036	Sulfur (%)	0693	0.29300	0.01200	0.31122	0.02175	0.01107	34	-0.84	036.42	0
036	Sulfur (%)	0171	0.29850	0.00100	0.31122	0.02175	0.01107	34	-0.59	036.42	0
036	Sulfur (%)	0187	0.29890	0.00280	0.31122	0.02175	0.01107	34	-0.57	036.42	0
036	Sulfur (%)	0035	0.30000	0.00000	0.31122	0.02175	0.01107	34	-0.52	036.42	0
036	Sulfur (%)	2089	0.30000	0.00000	0.31122	0.02175	0.01107	34	-0.52	036.43	0
036	Sulfur (%)	0692	0.30000	0.04000	0.31122	0.02175	0.01107	34	-0.52	036.42	0
036	Sulfur (%)	0229	0.30200	0.00800	0.31122	0.02175	0.01107	34	-0.42	036.04	0
036	Sulfur (%)	0345	0.30500	0.00400	0.31122	0.02175	0.01107	34	-0.29	036.43	0
036	Sulfur (%)	0407	0.30620	0.01740	0.31122	0.02175	0.01107	34	-0.23	036.42	0
036	Sulfur (%)	0889	0.30820	0.00460	0.31122	0.02175	0.01107	34	-0.14	036.99	0
036	Sulfur (%)	0297	0.31000	0.02000	0.31122	0.02175	0.01107	34	-0.06	036.43	0
036	Sulfur (%)	0629	0.31000	0.00000	0.31122	0.02175	0.01107	34	-0.06	036.43	0
036	Sulfur (%)	0918	0.31000	0.01600	0.31122	0.02175	0.01107	34	-0.06	036.43	0
036	Sulfur (%)	0560	0.31330	0.00940	0.31122	0.02175	0.01107	34	0.10	036.42	0
036	Sulfur (%)	0226	0.31500	0.03000	0.31122	0.02175	0.01107	34	0.17	036.04	0
036	Sulfur (%)	0278	0.32000	0.00000	0.31122	0.02175	0.01107	34	0.40	036.42	0
036	Sulfur (%)	0169	0.32000	0.00000	0.31122	0.02175	0.01107	34	0.40	036.43	0
036	Sulfur (%)	0242	0.32000	0.00000	0.31122	0.02175	0.01107	34	0.40	036.99	0
036	Sulfur (%)	0032	0.32105	0.01710	0.31122	0.02175	0.01107	34	0.45	036.43	0
036	Sulfur (%)	0098	0.32175	0.00050	0.31122	0.02175	0.01107	34	0.48	036.43	0
036	Sulfur (%)	0186	0.32550	0.00500	0.31122	0.02175	0.01107	34	0.66	036.52	0
036	Sulfur (%)	0357	0.32885	0.01870	0.31122	0.02175	0.01107	34	0.81	036.42	0
036	Sulfur (%)	0202	0.33650	0.00100	0.31122	0.02175	0.01107	34	1.16	036.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
036	Sulfur (%)	0508	0.33965	0.04990	0.31122	0.02175	0.01107	34	1.31	036.43	0
036	Sulfur (%)	0510	0.34000	0.00000	0.31122	0.02175	0.01107	34	1.32	036.43	0
036	Sulfur (%)	0294	0.34500	0.01000	0.31122	0.02175	0.01107	34	1.55	036.42	0
036	Sulfur (%)	0186	0.34900	0.01400	0.31122	0.02175	0.01107	34	1.74	036.42	0
036	Sulfur (%)	0037	0.37500	0.01000	0.31122	0.02175	0.01107	34	2.93	036.42	0
036	Sulfur (%)	0015	0.53000	0.02000	0.31122	0.02175	0.01107	34	10.06	036.04	2
037	Zinc (mg / kg (ppm))	2066	321.00	20.000	420.04	28.492	10.786	85	-3.48	037.53	0
037	Zinc (mg / kg (ppm))	2192	340.26	13.780	420.04	28.492	10.786	85	-2.80	037.43	0
037	Zinc (mg / kg (ppm))	0047	346.60	25.000	420.04	28.492	10.786	85	-2.58	037.52	0
037	Zinc (mg / kg (ppm))	0353	351.00	12.000	420.04	28.492	10.786	85	-2.42	037.43	0
037	Zinc (mg / kg (ppm))	2188	362.00	10.000	420.04	28.492	10.786	85	-2.04	037.31	0
037	Zinc (mg / kg (ppm))	0870	373.60	4.6000	420.04	28.492	10.786	85	-1.63	037.42	0
037	Zinc (mg / kg (ppm))	0169	375.50	5.0000	420.04	28.492	10.786	85	-1.56	037.43	0
037	Zinc (mg / kg (ppm))	0345	376.00	6.0000	420.04	28.492	10.786	85	-1.55	037.43	0
037	Zinc (mg / kg (ppm))	0186	378.00	22.000	420.04	28.492	10.786	85	-1.48	037.42	0
037	Zinc (mg / kg (ppm))	0034	382.25	13.100	420.04	28.492	10.786	85	-1.33	037.43	0
037	Zinc (mg / kg (ppm))	0955	383.00	16.000	420.04	28.492	10.786	85	-1.30	037.41	0
037	Zinc (mg / kg (ppm))	0692	383.50	11.000	420.04	28.492	10.786	85	-1.28	037.42	0
037	Zinc (mg / kg (ppm))	0045	385.50	5.0000	420.04	28.492	10.786	85	-1.21	037.42	0
037	Zinc (mg / kg (ppm))	0878	388.00	6.0000	420.04	28.492	10.786	85	-1.12	037.41	0
037	Zinc (mg / kg (ppm))	0049	389.95	26.000	420.04	28.492	10.786	85	-1.06	037.41	0
037	Zinc (mg / kg (ppm))	0263	390.46	8.8200	420.04	28.492	10.786	85	-1.04	037.42	0
037	Zinc (mg / kg (ppm))	2089	392.40	20.070	420.04	28.492	10.786	85	-0.97	037.43	0
037	Zinc (mg / kg (ppm))	0278	392.40	2.4000	420.04	28.492	10.786	85	-0.97	037.42	0
037	Zinc (mg / kg (ppm))	0043	393.00	12.000	420.04	28.492	10.786	85	-0.95	037.43	0
037	Zinc (mg / kg (ppm))	0889	393.69	1.8483	420.04	28.492	10.786	85	-0.92	037.99	0
037	Zinc (mg / kg (ppm))	0294	401.20	0.40000	420.04	28.492	10.786	85	-0.66	037.42	0
037	Zinc (mg / kg (ppm))	0171	402.50	7.0000	420.04	28.492	10.786	85	-0.62	037.41	0
037	Zinc (mg / kg (ppm))	0629	404.00	6.0000	420.04	28.492	10.786	85	-0.56	037.43	0
037	Zinc (mg / kg (ppm))	0629	404.50	11.000	420.04	28.492	10.786	85	-0.55	037.31	0
037	Zinc (mg / kg (ppm))	0918	404.71	12.310	420.04	28.492	10.786	85	-0.54	037.43	0
037	Zinc (mg / kg (ppm))	0520	406.50	5.0000	420.04	28.492	10.786	85	-0.48	037.41	0
037	Zinc (mg / kg (ppm))	0187	410.62	7.8400	420.04	28.492	10.786	85	-0.33	037.42	0
037	Zinc (mg / kg (ppm))	0098	412.15	15.100	420.04	28.492	10.786	85	-0.28	037.41	0
037	Zinc (mg / kg (ppm))	0508	412.84	23.650	420.04	28.492	10.786	85	-0.25	037.43	0
037	Zinc (mg / kg (ppm))	2145	416.10	4.2000	420.04	28.492	10.786	85	-0.14	037.41	0
037	Zinc (mg / kg (ppm))	0693	417.04	7.9160	420.04	28.492	10.786	85	-0.11	037.42	0
037	Zinc (mg / kg (ppm))	0874	417.50	15.000	420.04	28.492	10.786	85	-0.09	037.31	0
037	Zinc (mg / kg (ppm))	0037	417.60	15.400	420.04	28.492	10.786	85	-0.09	037.42	0
037	Zinc (mg / kg (ppm))	0682	418.28	0.00000	420.04	28.492	10.786	85	-0.06	037.43	0
037	Zinc (mg / kg (ppm))	0560	418.85	13.900	420.04	28.492	10.786	85	-0.04	037.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
037	Zinc (mg / kg (ppm))	0868	418.99	2.1600	420.04	28.492	10.786	85	-0.04	037.31	0
037	Zinc (mg / kg (ppm))	0511	419.50	15.000	420.04	28.492	10.786	85	-0.02	037.41	0
037	Zinc (mg / kg (ppm))	2195	420.57	3.3900	420.04	28.492	10.786	85	0.02	037.31	0
037	Zinc (mg / kg (ppm))	0019	421.55	4.5000	420.04	28.492	10.786	85	0.05	037.41	0
037	Zinc (mg / kg (ppm))	0032	422.35	6.7000	420.04	28.492	10.786	85	0.08	037.43	0
037	Zinc (mg / kg (ppm))	0675	423.61	5.7200	420.04	28.492	10.786	85	0.13	037.43	0
037	Zinc (mg / kg (ppm))	0014	424.00	6.0000	420.04	28.492	10.786	85	0.14	037.31	0
037	Zinc (mg / kg (ppm))	0098	424.45	3.5000	420.04	28.492	10.786	85	0.15	037.43	0
037	Zinc (mg / kg (ppm))	0723	425.00	0.00000	420.04	28.492	10.786	85	0.17	037.31	0
037	Zinc (mg / kg (ppm))	0010	425.00	10.000	420.04	28.492	10.786	85	0.17	037.33	0
037	Zinc (mg / kg (ppm))	0011	425.33	1.8100	420.04	28.492	10.786	85	0.19	037.41	0
037	Zinc (mg / kg (ppm))	0229	425.43	36.660	420.04	28.492	10.786	85	0.19	037.41	0
037	Zinc (mg / kg (ppm))	0598	425.90	12.000	420.04	28.492	10.786	85	0.21	037.41	0
037	Zinc (mg / kg (ppm))	0242	427.50	5.0000	420.04	28.492	10.786	85	0.26	037.99	0
037	Zinc (mg / kg (ppm))	0208	428.00	36.000	420.04	28.492	10.786	85	0.28	037.31	0
037	Zinc (mg / kg (ppm))	0001	428.55	9.7000	420.04	28.492	10.786	85	0.30	037.31	0
037	Zinc (mg / kg (ppm))	0042	429.00	20.000	420.04	28.492	10.786	85	0.31	037.42	0
037	Zinc (mg / kg (ppm))	0510	429.50	17.000	420.04	28.492	10.786	85	0.33	037.43	0
037	Zinc (mg / kg (ppm))	0939	430.07	0.23000	420.04	28.492	10.786	85	0.35	037.31	0
037	Zinc (mg / kg (ppm))	0968	430.50	1.0000	420.04	28.492	10.786	85	0.37	037.43	0
037	Zinc (mg / kg (ppm))	0407	430.85	1.8814	420.04	28.492	10.786	85	0.38	037.41	0
037	Zinc (mg / kg (ppm))	0202	432.38	2.1400	420.04	28.492	10.786	85	0.43	037.43	0
037	Zinc (mg / kg (ppm))	0910	433.00	24.000	420.04	28.492	10.786	85	0.45	037.41	0
037	Zinc (mg / kg (ppm))	0083	433.00	10.000	420.04	28.492	10.786	85	0.45	037.43	0
037	Zinc (mg / kg (ppm))	0948	433.62	1.1760	420.04	28.492	10.786	85	0.48	037.33	0
037	Zinc (mg / kg (ppm))	0354	434.40	10.270	420.04	28.492	10.786	85	0.50	037.31	0
037	Zinc (mg / kg (ppm))	2178	436.00	14.000	420.04	28.492	10.786	85	0.56	037.33	0
037	Zinc (mg / kg (ppm))	0003	436.00	10.000	420.04	28.492	10.786	85	0.56	037.41	0
037	Zinc (mg / kg (ppm))	0297	436.50	25.000	420.04	28.492	10.786	85	0.58	037.43	0
037	Zinc (mg / kg (ppm))	0689	437.00	5.8000	420.04	28.492	10.786	85	0.60	037.31	0
037	Zinc (mg / kg (ppm))	0357	437.73	18.587	420.04	28.492	10.786	85	0.62	037.42	0
037	Zinc (mg / kg (ppm))	0358	438.54	0.37000	420.04	28.492	10.786	85	0.65	037.41	0
037	Zinc (mg / kg (ppm))	0186	439.50	3.0000	420.04	28.492	10.786	85	0.68	037.52	0
037	Zinc (mg / kg (ppm))	0572	440.50	17.000	420.04	28.492	10.786	85	0.72	037.53	0
037	Zinc (mg / kg (ppm))	0590	440.50	11.000	420.04	28.492	10.786	85	0.72	037.99	0
037	Zinc (mg / kg (ppm))	0226	442.00	6.0000	420.04	28.492	10.786	85	0.77	037.41	0
037	Zinc (mg / kg (ppm))	2053	445.50	7.0000	420.04	28.492	10.786	85	0.89	037.42	0
037	Zinc (mg / kg (ppm))	0208	446.80	23.000	420.04	28.492	10.786	85	0.94	037.41	0
037	Zinc (mg / kg (ppm))	2196	447.00	0.00000	420.04	28.492	10.786	85	0.95	037.31	0
037	Zinc (mg / kg (ppm))	0413	450.00	28.000	420.04	28.492	10.786	85	1.05	037.42	0
037	Zinc (mg / kg (ppm))	0175	451.00	26.000	420.04	28.492	10.786	85	1.09	037.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
037	Zinc (mg / kg (ppm))	0190	454.06	1.6600	420.04	28.492	10.786	85	1.19	037.42	0
037	Zinc (mg / kg (ppm))	0726	455.35	5.5000	420.04	28.492	10.786	85	1.24	037.42	0
037	Zinc (mg / kg (ppm))	0512	459.30	9.8000	420.04	28.492	10.786	85	1.38	037.43	0
037	Zinc (mg / kg (ppm))	0100	460.00	30.000	420.04	28.492	10.786	85	1.40	037.99	0
037	Zinc (mg / kg (ppm))	0536	461.90	9.3000	420.04	28.492	10.786	85	1.47	037.31	0
037	Zinc (mg / kg (ppm))	0027	470.76	20.990	420.04	28.492	10.786	85	1.78	037.43	0
037	Zinc (mg / kg (ppm))	0425	474.65	6.5000	420.04	28.492	10.786	85	1.92	037.43	0
037	Zinc (mg / kg (ppm))	0609	477.50	5.0000	420.04	28.492	10.786	85	2.02	037.32	0
037	Zinc (mg / kg (ppm))	0405	546.85	2.1000	420.04	28.492	10.786	85	4.45	037.41	0
038	Molybdenum (mg / kg (ppm))	0345	2.4350	0.03000	2.7957	0.27134	0.21103	15	-1.33	038.43	0
038	Molybdenum (mg / kg (ppm))	0353	2.5196	0.34340	2.7957	0.27134	0.21103	15	-1.02	038.43	0
038	Molybdenum (mg / kg (ppm))	0171	2.5500	0.00000	2.7957	0.27134	0.21103	15	-0.91	038.41	0
038	Molybdenum (mg / kg (ppm))	0278	2.5650	0.15000	2.7957	0.27134	0.21103	15	-0.85	038.42	0
038	Molybdenum (mg / kg (ppm))	0226	2.5950	0.03000	2.7957	0.27134	0.21103	15	-0.74	038.41	0
038	Molybdenum (mg / kg (ppm))	0098	2.6320	0.18800	2.7957	0.27134	0.21103	15	-0.60	038.53	0
038	Molybdenum (mg / kg (ppm))	0910	2.7700	0.30000	2.7957	0.27134	0.21103	15	-0.09	038.52	0
038	Molybdenum (mg / kg (ppm))	0918	2.7958	0.31630	2.7957	0.27134	0.21103	15	0.00	038.53	0
038	Molybdenum (mg / kg (ppm))	0169	2.8000	0.12000	2.7957	0.27134	0.21103	15	0.02	038.43	0
038	Molybdenum (mg / kg (ppm))	0508	2.9826	0.13880	2.7957	0.27134	0.21103	15	0.69	038.43	0
038	Molybdenum (mg / kg (ppm))	0693	2.9885	0.34900	2.7957	0.27134	0.21103	15	0.71	038.42	0
038	Molybdenum (mg / kg (ppm))	0297	3.0000	0.76000	2.7957	0.27134	0.21103	15	0.75	038.43	0
038	Molybdenum (mg / kg (ppm))	0186	3.0000	0.00000	2.7957	0.27134	0.21103	15	0.75	038.52	0
038	Molybdenum (mg / kg (ppm))	0510	3.1000	0.00000	2.7957	0.27134	0.21103	15	1.12	038.43	0
038	Molybdenum (mg / kg (ppm))	0045	3.2900	0.44000	2.7957	0.27134	0.21103	15	1.82	038.42	0
040	Barium (mg / kg (ppm))	0918	21.796	0.66140				3		040.53	0
040	Barium (mg / kg (ppm))	0560	23.615	0.05000				3		040.42	0
040	Barium (mg / kg (ppm))	0508	23.783	0.53600				3		040.43	0
041	Vanadium (mg / kg (ppm))	0508	2.6696	0.18080				1		041.43	0
042	Chloride (%)	0948	0.72950	0.00100				3		042.00	0
042	Chloride (%)	2188	0.74000	0.00000				3		042.01	0
042	Chloride (%)	0169	0.76000	0.00000				3		042.00	0
042	Chloride (%)	0297	0.77500	0.03000				3		042.00	1
101	Choline Chloride (mg / kg (ppm))	0939	1,429.0	90.000				1		101.01	0
103	Pantothenic Acid (mg / kg (ppm))	0629	24.450	0.50000				1		103.02	0
104	Riboflavin (mg / kg (ppm))	0910	13.005	0.11000				3		104.03	0
104	Riboflavin (mg / kg (ppm))	2192	15.950	1.5000				3		104.03	0
104	Riboflavin (mg / kg (ppm))	0171	20.500	1.0000				3		104.00	0
105	Thiamine (mg / kg (ppm))	0629	8.0450	0.01000				2		105.00	0
105	Thiamine (mg / kg (ppm))	0910	20.755	0.23000				2		105.00	0
106	Vitamin A (KU / kg)	0875	5.1800	1.3400	10.809	4.5191	0.93644	18	-1.25	106.02	0
106	Vitamin A (KU / kg)	0910	6.4000	1.8000	10.809	4.5191	0.93644	18	-0.98	106.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
106	Vitamin A (KU / kg)	0004	7.2300	1.0600	10.809	4.5191	0.93644	18	-0.79	106.02	0
106	Vitamin A (KU / kg)	0689	7.9000	0.20000	10.809	4.5191	0.93644	18	-0.64	106.02	0
106	Vitamin A (KU / kg)	0610	8.1500	0.30000	10.809	4.5191	0.93644	18	-0.59	106.02	0
106	Vitamin A (KU / kg)	2053	9.4500	1.1000	10.809	4.5191	0.93644	18	-0.30	106.02	0
106	Vitamin A (KU / kg)	0171	9.5750	0.65000	10.809	4.5191	0.93644	18	-0.27	106.00	0
106	Vitamin A (KU / kg)	0512	9.6270	0.16400	10.809	4.5191	0.93644	18	-0.26	106.02	0
106	Vitamin A (KU / kg)	0003	9.8700	0.72000	10.809	4.5191	0.93644	18	-0.21	106.02	0
106	Vitamin A (KU / kg)	0918	9.8900	0.22000	10.809	4.5191	0.93644	18	-0.20	106.02	0
106	Vitamin A (KU / kg)	2103	9.9050	0.95000	10.809	4.5191	0.93644	18	-0.20	106.02	0
106	Vitamin A (KU / kg)	0676	10.350	0.12000	10.809	4.5191	0.93644	18	-0.10	106.02	0
106	Vitamin A (KU / kg)	0208	10.384	0.35300	10.809	4.5191	0.93644	18	-0.09	106.02	0
106	Vitamin A (KU / kg)	0098	10.395	4.0100	10.809	4.5191	0.93644	18	-0.09	106.01	0
106	Vitamin A (KU / kg)	0169	17.500	1.0000	10.809	4.5191	0.93644	18	1.48	106.02	0
106	Vitamin A (KU / kg)	0905	20.627	0.16900	10.809	4.5191	0.93644	18	2.17	106.02	0
106	Vitamin A (KU / kg)	2066	25.750	2.7000	10.809	4.5191	0.93644	18	3.31	106.02	0
106	Vitamin A (KU / kg)	2196	50.390	0.00000	10.809	4.5191	0.93644	18	8.76	106.02	0
106	Vitamin A (KU / kg)	0675	24.565	6.1700	10.809	4.5191	0.93644	18	3.04	106.02	1
108	Vitamin D3 (KU / kg)	0910	0.49500	0.05000	2.9456	2.8998	0.26000	5		108.02	0
108	Vitamin D3 (KU / kg)	0918	0.95050	0.05500	2.9456	2.8998	0.26000	5		108.02	0
108	Vitamin D3 (KU / kg)	0208	1.2525	0.65500	2.9456	2.8998	0.26000	5		108.02	0
108	Vitamin D3 (KU / kg)	0675	5.0300	0.54000	2.9456	2.8998	0.26000	5		108.02	0
108	Vitamin D3 (KU / kg)	0169	7.0000	0.00000	2.9456	2.8998	0.26000	5		108.01	0
109	Vitamin E (IU/kg)	2053	485.00	6.0000	573.43	64.176	25.278	19	-1.38	109.02	0
109	Vitamin E (IU/kg)	0675	490.96	11.240	573.43	64.176	25.278	19	-1.28	109.02	0
109	Vitamin E (IU/kg)	0610	508.50	1.0000	573.43	64.176	25.278	19	-1.01	109.02	0
109	Vitamin E (IU/kg)	0590	517.00	28.000	573.43	64.176	25.278	19	-0.88	109.02	0
109	Vitamin E (IU/kg)	2103	520.00	42.000	573.43	64.176	25.278	19	-0.83	109.02	0
109	Vitamin E (IU/kg)	0955	536.50	87.000	573.43	64.176	25.278	19	-0.58	109.02	0
109	Vitamin E (IU/kg)	2066	540.50	75.000	573.43	64.176	25.278	19	-0.51	109.02	0
109	Vitamin E (IU/kg)	0171	559.50	11.000	573.43	64.176	25.278	19	-0.22	109.99	0
109	Vitamin E (IU/kg)	0676	564.48	5.1300	573.43	64.176	25.278	19	-0.14	109.02	0
109	Vitamin E (IU/kg)	2195	572.00	14.000	573.43	64.176	25.278	19	-0.02	109.02	0
109	Vitamin E (IU/kg)	0208	580.35	58.309	573.43	64.176	25.278	19	0.11	109.02	0
109	Vitamin E (IU/kg)	0098	586.50	9.0000	573.43	64.176	25.278	19	0.20	109.02	0
109	Vitamin E (IU/kg)	0723	601.39	0.00000	573.43	64.176	25.278	19	0.44	109.02	0
109	Vitamin E (IU/kg)	2196	605.18	0.00000	573.43	64.176	25.278	19	0.49	109.02	0
109	Vitamin E (IU/kg)	0918	620.50	51.000	573.43	64.176	25.278	19	0.73	109.02	0
109	Vitamin E (IU/kg)	0910	629.60	23.000	573.43	64.176	25.278	19	0.88	109.02	0
109	Vitamin E (IU/kg)	0027	637.75	2.6020	573.43	64.176	25.278	19	1.00	109.02	0
109	Vitamin E (IU/kg)	0905	730.00	50.000	573.43	64.176	25.278	19	2.44	109.02	0
109	Vitamin E (IU/kg)	0169	757.00	6.0000	573.43	64.176	25.278	19	2.86	109.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
111	Vitamin C, phosphorylated (mg / kg (ppn	0918	48.700	1.2000				3		111.01	0
111	Vitamin C, phosphorylated (mg / kg (ppn	2066	138.50	5.0000				3		111.01	0
111	Vitamin C, phosphorylated (mg / kg (ppn	2076	197.00	78.000				3		111.00	0
120	Alanine (%)	0939	0.71000	0.00000	0.85019	0.02631	0.00695	22	-5.33	120.05	0
120	Alanine (%)	0910	0.79000	0.00000	0.85019	0.02631	0.00695	22	-2.29	120.00	0
120	Alanine (%)	0652	0.82000	0.00000	0.85019	0.02631	0.00695	22	-1.15	120.00	0
120	Alanine (%)	0859	0.82500	0.01680	0.85019	0.02631	0.00695	22	-0.96	120.00	0
120	Alanine (%)	0968	0.83000	0.00000	0.85019	0.02631	0.00695	22	-0.77	120.00	0
120	Alanine (%)	0870	0.83670	0.00620	0.85019	0.02631	0.00695	22	-0.51	120.00	0
120	Alanine (%)	0878	0.83950	0.00900	0.85019	0.02631	0.00695	22	-0.41	120.00	0
120	Alanine (%)	0354	0.84400	0.00200	0.85019	0.02631	0.00695	22	-0.24	120.00	0
120	Alanine (%)	0571	0.84400	0.00800	0.85019	0.02631	0.00695	22	-0.24	120.00	0
120	Alanine (%)	0872	0.84400	0.01000	0.85019	0.02631	0.00695	22	-0.24	120.00	0
120	Alanine (%)	2059	0.84600	0.00000	0.85019	0.02631	0.00695	22	-0.16	120.00	0
120	Alanine (%)	0644	0.85000	0.01200	0.85019	0.02631	0.00695	22	-0.01	120.00	0
120	Alanine (%)	0918	0.85280	0.01600	0.85019	0.02631	0.00695	22	0.10	120.00	0
120	Alanine (%)	2196	0.85600	0.00000	0.85019	0.02631	0.00695	22	0.22	120.00	0
120	Alanine (%)	0098	0.85650	0.01700	0.85019	0.02631	0.00695	22	0.24	120.02	0
120	Alanine (%)	0504	0.86000	0.02000	0.85019	0.02631	0.00695	22	0.37	120.00	0
120	Alanine (%)	0675	0.86500	0.01000	0.85019	0.02631	0.00695	22	0.56	120.00	0
120	Alanine (%)	0676	0.87500	0.01000	0.85019	0.02631	0.00695	22	0.94	120.05	0
120	Alanine (%)	0171	0.87750	0.00100	0.85019	0.02631	0.00695	22	1.04	120.00	0
120	Alanine (%)	2188	0.88500	0.01000	0.85019	0.02631	0.00695	22	1.32	120.05	0
120	Alanine (%)	0868	0.88600	0.00200	0.85019	0.02631	0.00695	22	1.36	120.00	0
120	Alanine (%)	0684	1.0375	0.00300	0.85019	0.02631	0.00695	22	7.12	120.00	0
120	Alanine (%)	0941	0.88250	0.04300	0.85019	0.02631	0.00695	22	1.23	120.00	1
121	Arginine (%)	0870	1.0311	0.00600	1.1106	0.04217	0.01819	22	-1.89	121.00	0
121	Arginine (%)	0968	1.0495	0.00100	1.1106	0.04217	0.01819	22	-1.45	121.00	0
121	Arginine (%)	0859	1.0503	0.01960	1.1106	0.04217	0.01819	22	-1.43	121.00	0
121	Arginine (%)	0878	1.0650	0.01800	1.1106	0.04217	0.01819	22	-1.08	121.00	0
121	Arginine (%)	0504	1.0850	0.01000	1.1106	0.04217	0.01819	22	-0.61	121.00	0
121	Arginine (%)	0652	1.0850	0.05000	1.1106	0.04217	0.01819	22	-0.61	121.00	0
121	Arginine (%)	0171	1.0855	0.00100	1.1106	0.04217	0.01819	22	-0.60	121.00	0
121	Arginine (%)	0354	1.0925	0.00100	1.1106	0.04217	0.01819	22	-0.43	121.00	0
121	Arginine (%)	0098	1.0990	0.01800	1.1106	0.04217	0.01819	22	-0.28	121.02	0
121	Arginine (%)	0868	1.1090	0.00800	1.1106	0.04217	0.01819	22	-0.04	121.00	0
121	Arginine (%)	0684	1.1125	0.00300	1.1106	0.04217	0.01819	22	0.04	121.00	0
121	Arginine (%)	2059	1.1195	0.00500	1.1106	0.04217	0.01819	22	0.21	121.00	0
121	Arginine (%)	0918	1.1223	0.03650	1.1106	0.04217	0.01819	22	0.28	121.00	0
121	Arginine (%)	0872	1.1295	0.00700	1.1106	0.04217	0.01819	22	0.45	121.00	0
121	Arginine (%)	0910	1.1300	0.08000	1.1106	0.04217	0.01819	22	0.46	121.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
121	Arginine (%)	0644	1.1335	0.01500	1.1106	0.04217	0.01819	22	0.54	121.00	0
121	Arginine (%)	0675	1.1350	0.01000	1.1106	0.04217	0.01819	22	0.58	121.00	0
121	Arginine (%)	0571	1.1455	0.00900	1.1106	0.04217	0.01819	22	0.83	121.00	0
121	Arginine (%)	2196	1.1470	0.00000	1.1106	0.04217	0.01819	22	0.86	121.00	0
121	Arginine (%)	2188	1.1550	0.01000	1.1106	0.04217	0.01819	22	1.05	121.05	0
121	Arginine (%)	0676	1.1650	0.03000	1.1106	0.04217	0.01819	22	1.29	121.05	0
121	Arginine (%)	0941	1.1710	0.06200	1.1106	0.04217	0.01819	22	1.43	121.00	0
121	Arginine (%)	0939	0.99500	0.15000	1.1106	0.04217	0.01819	22	-2.74	121.05	1
122	Aspartic (%)	0939	1.6500	0.10000	1.7878	0.05672	0.03209	23	-2.43	122.05	0
122	Aspartic (%)	0859	1.6533	0.03180	1.7878	0.05672	0.03209	23	-2.37	122.00	0
122	Aspartic (%)	0870	1.7277	0.05600	1.7878	0.05672	0.03209	23	-1.06	122.00	0
122	Aspartic (%)	0910	1.7450	0.03000	1.7878	0.05672	0.03209	23	-0.76	122.00	0
122	Aspartic (%)	0968	1.7465	0.00700	1.7878	0.05672	0.03209	23	-0.73	122.00	0
122	Aspartic (%)	0504	1.7500	0.06000	1.7878	0.05672	0.03209	23	-0.67	122.00	0
122	Aspartic (%)	0872	1.7570	0.00400	1.7878	0.05672	0.03209	23	-0.54	122.00	0
122	Aspartic (%)	2059	1.7585	0.00700	1.7878	0.05672	0.03209	23	-0.52	122.00	0
122	Aspartic (%)	0878	1.7710	0.01200	1.7878	0.05672	0.03209	23	-0.30	122.00	0
122	Aspartic (%)	0354	1.7740	0.00800	1.7878	0.05672	0.03209	23	-0.24	122.00	0
122	Aspartic (%)	0571	1.7750	0.01600	1.7878	0.05672	0.03209	23	-0.23	122.00	0
122	Aspartic (%)	0652	1.7750	0.03000	1.7878	0.05672	0.03209	23	-0.23	122.00	0
122	Aspartic (%)	0644	1.7825	0.00700	1.7878	0.05672	0.03209	23	-0.09	122.00	0
122	Aspartic (%)	0941	1.8030	0.11400	1.7878	0.05672	0.03209	23	0.27	122.00	0
122	Aspartic (%)	0684	1.8055	0.02500	1.7878	0.05672	0.03209	23	0.31	122.00	0
122	Aspartic (%)	0098	1.8120	0.02400	1.7878	0.05672	0.03209	23	0.43	122.02	0
122	Aspartic (%)	0675	1.8150	0.01000	1.7878	0.05672	0.03209	23	0.48	122.00	0
122	Aspartic (%)	2196	1.8210	0.00000	1.7878	0.05672	0.03209	23	0.58	122.00	0
122	Aspartic (%)	2188	1.8350	0.07000	1.7878	0.05672	0.03209	23	0.83	122.05	0
122	Aspartic (%)	0868	1.8570	0.01200	1.7878	0.05672	0.03209	23	1.22	122.00	0
122	Aspartic (%)	0171	1.8620	0.00400	1.7878	0.05672	0.03209	23	1.31	122.00	0
122	Aspartic (%)	0918	1.8692	0.02030	1.7878	0.05672	0.03209	23	1.43	122.00	0
122	Aspartic (%)	0676	1.8750	0.09000	1.7878	0.05672	0.03209	23	1.54	122.05	0
124	Cysteine/Cystine (%)	0870	0.26255	0.00350	0.30168	0.02246	0.00507	24	-1.74	124.00	0
124	Cysteine/Cystine (%)	0918	0.26500	0.01000	0.30168	0.02246	0.00507	24	-1.63	124.00	0
124	Cysteine/Cystine (%)	0939	0.26500	0.01000	0.30168	0.02246	0.00507	24	-1.63	124.05	0
124	Cysteine/Cystine (%)	0868	0.27950	0.00700	0.30168	0.02246	0.00507	24	-0.99	124.00	0
124	Cysteine/Cystine (%)	0652	0.28500	0.01000	0.30168	0.02246	0.00507	24	-0.74	124.00	0
124	Cysteine/Cystine (%)	0644	0.28800	0.00000	0.30168	0.02246	0.00507	24	-0.61	124.00	0
124	Cysteine/Cystine (%)	0910	0.29000	0.02000	0.30168	0.02246	0.00507	24	-0.52	124.00	0
124	Cysteine/Cystine (%)	0872	0.29400	0.00200	0.30168	0.02246	0.00507	24	-0.34	124.00	0
124	Cysteine/Cystine (%)	2196	0.29800	0.00000	0.30168	0.02246	0.00507	24	-0.16	124.00	0
124	Cysteine/Cystine (%)	0571	0.29950	0.00300	0.30168	0.02246	0.00507	24	-0.10	124.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
124	Cysteine/Cystine (%)	0684	0.29950	0.00100	0.30168	0.02246	0.00507	24	-0.10	124.00	0
124	Cysteine/Cystine (%)	0968	0.30000	0.00000	0.30168	0.02246	0.00507	24	-0.07	124.00	0
124	Cysteine/Cystine (%)	0878	0.30100	0.00800	0.30168	0.02246	0.00507	24	-0.03	124.00	0
124	Cysteine/Cystine (%)	0941	0.30100	0.00000	0.30168	0.02246	0.00507	24	-0.03	124.00	0
124	Cysteine/Cystine (%)	0723	0.30760	0.00000	0.30168	0.02246	0.00507	24	0.26	124.05	0
124	Cysteine/Cystine (%)	0171	0.31000	0.00200	0.30168	0.02246	0.00507	24	0.37	124.00	0
124	Cysteine/Cystine (%)	2059	0.31100	0.00600	0.30168	0.02246	0.00507	24	0.42	124.00	0
124	Cysteine/Cystine (%)	0859	0.31110	0.00320	0.30168	0.02246	0.00507	24	0.42	124.00	0
124	Cysteine/Cystine (%)	0098	0.31400	0.02400	0.30168	0.02246	0.00507	24	0.55	124.02	0
124	Cysteine/Cystine (%)	0504	0.32000	0.00000	0.30168	0.02246	0.00507	24	0.82	124.00	0
124	Cysteine/Cystine (%)	0354	0.32100	0.00200	0.30168	0.02246	0.00507	24	0.86	124.00	0
124	Cysteine/Cystine (%)	2195	0.35000	0.00000	0.30168	0.02246	0.00507	24	2.15	124.05	0
124	Cysteine/Cystine (%)	0675	0.35500	0.01000	0.30168	0.02246	0.00507	24	2.37	124.00	0
124	Cysteine/Cystine (%)	2188	0.38000	0.00000	0.30168	0.02246	0.00507	24	3.49	124.05	0
125	Glutamic (%)	0171	2.6405	0.00300	2.9907	0.08729	0.04523	23	-4.01	125.00	0
125	Glutamic (%)	0859	2.8629	0.05970	2.9907	0.08729	0.04523	23	-1.46	125.00	0
125	Glutamic (%)	2059	2.9030	0.01000	2.9907	0.08729	0.04523	23	-1.00	125.00	0
125	Glutamic (%)	0878	2.9165	0.03300	2.9907	0.08729	0.04523	23	-0.85	125.00	0
125	Glutamic (%)	0872	2.9345	0.00300	2.9907	0.08729	0.04523	23	-0.64	125.00	0
125	Glutamic (%)	0504	2.9350	0.07000	2.9907	0.08729	0.04523	23	-0.64	125.00	0
125	Glutamic (%)	2196	2.9390	0.00000	2.9907	0.08729	0.04523	23	-0.59	125.00	0
125	Glutamic (%)	0571	2.9460	0.03600	2.9907	0.08729	0.04523	23	-0.51	125.00	0
125	Glutamic (%)	0868	2.9505	0.01500	2.9907	0.08729	0.04523	23	-0.46	125.00	0
125	Glutamic (%)	0354	2.9710	0.00400	2.9907	0.08729	0.04523	23	-0.23	125.00	0
125	Glutamic (%)	0968	2.9745	0.01100	2.9907	0.08729	0.04523	23	-0.19	125.00	0
125	Glutamic (%)	0644	2.9775	0.03500	2.9907	0.08729	0.04523	23	-0.15	125.00	0
125	Glutamic (%)	0652	2.9850	0.03000	2.9907	0.08729	0.04523	23	-0.06	125.00	0
125	Glutamic (%)	0098	2.9930	0.06600	2.9907	0.08729	0.04523	23	0.03	125.02	0
125	Glutamic (%)	0910	3.0200	0.12000	2.9907	0.08729	0.04523	23	0.34	125.00	0
125	Glutamic (%)	0939	3.0300	0.18000	2.9907	0.08729	0.04523	23	0.45	125.05	0
125	Glutamic (%)	2188	3.0300	0.06000	2.9907	0.08729	0.04523	23	0.45	125.05	0
125	Glutamic (%)	0870	3.0493	0.04170	2.9907	0.08729	0.04523	23	0.67	125.00	0
125	Glutamic (%)	0918	3.0826	0.00400	2.9907	0.08729	0.04523	23	1.05	125.00	0
125	Glutamic (%)	0941	3.0870	0.09400	2.9907	0.08729	0.04523	23	1.10	125.00	0
125	Glutamic (%)	0676	3.0950	0.09000	2.9907	0.08729	0.04523	23	1.20	125.05	0
125	Glutamic (%)	0675	3.1550	0.05000	2.9907	0.08729	0.04523	23	1.88	125.00	0
125	Glutamic (%)	0684	3.2085	0.02500	2.9907	0.08729	0.04523	23	2.50	125.00	0
126	Glycine (%)	0939	0.75500	0.05000	0.84682	0.03919	0.01200	23	-2.34	126.05	0
126	Glycine (%)	0870	0.79530	0.00220	0.84682	0.03919	0.01200	23	-1.31	126.00	0
126	Glycine (%)	0859	0.79885	0.01610	0.84682	0.03919	0.01200	23	-1.22	126.00	0
126	Glycine (%)	0171	0.80300	0.00000	0.84682	0.03919	0.01200	23	-1.12	126.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
126	Glycine (%)	0652	0.81000	0.00000	0.84682	0.03919	0.01200	23	-0.94	126.00	0
126	Glycine (%)	0910	0.82000	0.00000	0.84682	0.03919	0.01200	23	-0.68	126.00	0
126	Glycine (%)	2188	0.83000	0.06000	0.84682	0.03919	0.01200	23	-0.43	126.05	0
126	Glycine (%)	0354	0.83100	0.00000	0.84682	0.03919	0.01200	23	-0.40	126.00	0
126	Glycine (%)	2059	0.83800	0.00400	0.84682	0.03919	0.01200	23	-0.23	126.00	0
126	Glycine (%)	0878	0.83850	0.00700	0.84682	0.03919	0.01200	23	-0.21	126.00	0
126	Glycine (%)	0968	0.84000	0.00000	0.84682	0.03919	0.01200	23	-0.17	126.00	0
126	Glycine (%)	0872	0.84600	0.00200	0.84682	0.03919	0.01200	23	-0.02	126.00	0
126	Glycine (%)	0571	0.84900	0.00800	0.84682	0.03919	0.01200	23	0.06	126.00	0
126	Glycine (%)	0504	0.85500	0.01000	0.84682	0.03919	0.01200	23	0.21	126.00	0
126	Glycine (%)	0675	0.86000	0.00000	0.84682	0.03919	0.01200	23	0.34	126.00	0
126	Glycine (%)	0644	0.86050	0.00500	0.84682	0.03919	0.01200	23	0.35	126.00	0
126	Glycine (%)	0098	0.86500	0.02200	0.84682	0.03919	0.01200	23	0.46	126.02	0
126	Glycine (%)	0941	0.87950	0.05300	0.84682	0.03919	0.01200	23	0.83	126.00	0
126	Glycine (%)	0918	0.88560	0.02260	0.84682	0.03919	0.01200	23	0.99	126.00	0
126	Glycine (%)	2196	0.88800	0.00000	0.84682	0.03919	0.01200	23	1.05	126.00	0
126	Glycine (%)	0684	0.89200	0.01200	0.84682	0.03919	0.01200	23	1.15	126.00	0
126	Glycine (%)	0868	0.89800	0.00200	0.84682	0.03919	0.01200	23	1.31	126.00	0
126	Glycine (%)	0676	0.91000	0.00000	0.84682	0.03919	0.01200	23	1.61	126.05	0
127	Histidine (%)	0939	0.36500	0.03000	0.41934	0.02108	0.00683	22	-2.58	127.05	0
127	Histidine (%)	0859	0.37480	0.00680	0.41934	0.02108	0.00683	22	-2.11	127.00	0
127	Histidine (%)	0171	0.39500	0.00400	0.41934	0.02108	0.00683	22	-1.15	127.00	0
127	Histidine (%)	0870	0.39800	0.00840	0.41934	0.02108	0.00683	22	-1.01	127.00	0
127	Histidine (%)	2059	0.40300	0.00600	0.41934	0.02108	0.00683	22	-0.78	127.00	0
127	Histidine (%)	2188	0.40500	0.01000	0.41934	0.02108	0.00683	22	-0.68	127.05	0
127	Histidine (%)	0676	0.41000	0.00000	0.41934	0.02108	0.00683	22	-0.44	127.05	0
127	Histidine (%)	0098	0.41550	0.00500	0.41934	0.02108	0.00683	22	-0.18	127.02	0
127	Histidine (%)	0571	0.41700	0.00400	0.41934	0.02108	0.00683	22	-0.11	127.00	0
127	Histidine (%)	0968	0.41850	0.00300	0.41934	0.02108	0.00683	22	-0.04	127.00	0
127	Histidine (%)	0354	0.41900	0.00000	0.41934	0.02108	0.00683	22	-0.02	127.00	0
127	Histidine (%)	0684	0.41900	0.00200	0.41934	0.02108	0.00683	22	-0.02	127.00	0
127	Histidine (%)	0918	0.41955	0.00610	0.41934	0.02108	0.00683	22	0.01	127.00	0
127	Histidine (%)	0910	0.42000	0.02000	0.41934	0.02108	0.00683	22	0.03	127.00	0
127	Histidine (%)	0872	0.42700	0.00400	0.41934	0.02108	0.00683	22	0.36	127.00	0
127	Histidine (%)	0675	0.43000	0.00000	0.41934	0.02108	0.00683	22	0.51	127.00	0
127	Histidine (%)	0868	0.43100	0.01000	0.41934	0.02108	0.00683	22	0.55	127.00	0
127	Histidine (%)	0644	0.43650	0.00300	0.41934	0.02108	0.00683	22	0.81	127.00	0
127	Histidine (%)	0941	0.43950	0.01300	0.41934	0.02108	0.00683	22	0.96	127.00	0
127	Histidine (%)	0878	0.44450	0.00500	0.41934	0.02108	0.00683	22	1.19	127.00	0
127	Histidine (%)	0504	0.46500	0.01000	0.41934	0.02108	0.00683	22	2.17	127.00	0
127	Histidine (%)	2196	0.46600	0.00000	0.41934	0.02108	0.00683	22	2.21	127.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
127	Histidine (%)	0652	0.43000	0.04000	0.41934	0.02108	0.00683	22	0.51	127.00	1
128	Isoleucine (%)	0939	0.63000	0.04000	0.71763	0.04034	0.01196	23	-2.17	128.05	0
128	Isoleucine (%)	0652	0.66000	0.00000	0.71763	0.04034	0.01196	23	-1.43	128.00	0
128	Isoleucine (%)	0918	0.66680	0.02260	0.71763	0.04034	0.01196	23	-1.26	128.00	0
128	Isoleucine (%)	0872	0.67900	0.00400	0.71763	0.04034	0.01196	23	-0.96	128.00	0
128	Isoleucine (%)	0870	0.68395	0.03230	0.71763	0.04034	0.01196	23	-0.83	128.00	0
128	Isoleucine (%)	0684	0.68450	0.00300	0.71763	0.04034	0.01196	23	-0.82	128.00	0
128	Isoleucine (%)	0878	0.69300	0.02000	0.71763	0.04034	0.01196	23	-0.61	128.00	0
128	Isoleucine (%)	0354	0.70100	0.00200	0.71763	0.04034	0.01196	23	-0.41	128.00	0
128	Isoleucine (%)	0676	0.70500	0.01000	0.71763	0.04034	0.01196	23	-0.31	128.05	0
128	Isoleucine (%)	0968	0.70750	0.01500	0.71763	0.04034	0.01196	23	-0.25	128.00	0
128	Isoleucine (%)	2059	0.71650	0.00700	0.71763	0.04034	0.01196	23	-0.03	128.00	0
128	Isoleucine (%)	0910	0.72000	0.00000	0.71763	0.04034	0.01196	23	0.06	128.00	0
128	Isoleucine (%)	0868	0.72550	0.01700	0.71763	0.04034	0.01196	23	0.20	128.00	0
128	Isoleucine (%)	0571	0.72850	0.01300	0.71763	0.04034	0.01196	23	0.27	128.00	0
128	Isoleucine (%)	0098	0.72950	0.00500	0.71763	0.04034	0.01196	23	0.29	128.02	0
128	Isoleucine (%)	0859	0.73190	0.01320	0.71763	0.04034	0.01196	23	0.35	128.00	0
128	Isoleucine (%)	2196	0.74700	0.00000	0.71763	0.04034	0.01196	23	0.73	128.00	0
128	Isoleucine (%)	0644	0.75050	0.00300	0.71763	0.04034	0.01196	23	0.81	128.00	0
128	Isoleucine (%)	0171	0.75250	0.00300	0.71763	0.04034	0.01196	23	0.86	128.00	0
128	Isoleucine (%)	2188	0.76000	0.02000	0.71763	0.04034	0.01196	23	1.05	128.05	0
128	Isoleucine (%)	0941	0.76250	0.02500	0.71763	0.04034	0.01196	23	1.11	128.00	0
128	Isoleucine (%)	0504	0.76500	0.01000	0.71763	0.04034	0.01196	23	1.17	128.00	0
128	Isoleucine (%)	0675	0.79500	0.01000	0.71763	0.04034	0.01196	23	1.92	128.00	0
129	Leucine (%)	0939	1.0950	0.15000	1.2605	0.04748	0.01806	22	-3.49	129.05	0
129	Leucine (%)	0652	1.2000	0.02000	1.2605	0.04748	0.01806	22	-1.28	129.00	0
129	Leucine (%)	0968	1.2000	0.00000	1.2605	0.04748	0.01806	22	-1.28	129.00	0
129	Leucine (%)	0870	1.2071	0.00210	1.2605	0.04748	0.01806	22	-1.13	129.00	0
129	Leucine (%)	0878	1.2300	0.01200	1.2605	0.04748	0.01806	22	-0.64	129.00	0
129	Leucine (%)	0872	1.2325	0.00100	1.2605	0.04748	0.01806	22	-0.59	129.00	0
129	Leucine (%)	0354	1.2345	0.00300	1.2605	0.04748	0.01806	22	-0.55	129.00	0
129	Leucine (%)	2059	1.2465	0.00100	1.2605	0.04748	0.01806	22	-0.30	129.00	0
129	Leucine (%)	2196	1.2480	0.00000	1.2605	0.04748	0.01806	22	-0.26	129.00	0
129	Leucine (%)	0859	1.2505	0.02430	1.2605	0.04748	0.01806	22	-0.21	129.00	0
129	Leucine (%)	0910	1.2550	0.03000	1.2605	0.04748	0.01806	22	-0.12	129.00	0
129	Leucine (%)	0676	1.2550	0.01000	1.2605	0.04748	0.01806	22	-0.12	129.05	0
129	Leucine (%)	0571	1.2695	0.01900	1.2605	0.04748	0.01806	22	0.19	129.00	0
129	Leucine (%)	0684	1.2705	0.01100	1.2605	0.04748	0.01806	22	0.21	129.00	0
129	Leucine (%)	0171	1.2840	0.00000	1.2605	0.04748	0.01806	22	0.49	129.00	0
129	Leucine (%)	0644	1.2870	0.00600	1.2605	0.04748	0.01806	22	0.56	129.00	0
129	Leucine (%)	0098	1.2950	0.02600	1.2605	0.04748	0.01806	22	0.73	129.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
129	Leucine (%)	0504	1.3000	0.00000	1.2605	0.04748	0.01806	22	0.83	129.00	0
129	Leucine (%)	0918	1.3060	0.00290	1.2605	0.04748	0.01806	22	0.96	129.00	0
129	Leucine (%)	0941	1.3080	0.06600	1.2605	0.04748	0.01806	22	1.00	129.00	0
129	Leucine (%)	0868	1.3425	0.00300	1.2605	0.04748	0.01806	22	1.73	129.00	0
129	Leucine (%)	0675	1.7750	0.01000	1.2605	0.04748	0.01806	22	10.84	129.00	0
129	Leucine (%)	2188	65.155	127.69	1.2605	0.04748	0.01806	22	1345.81	129.05	2
130	L-Lysine (%)	0171	0.81100	0.00200	0.97656	0.05171	0.01674	29	-3.20	130.00	0
130	L-Lysine (%)	2053	0.81950	0.03700	0.97656	0.05171	0.01674	29	-3.04	130.99	0
130	L-Lysine (%)	0848	0.89000	0.08000	0.97656	0.05171	0.01674	29	-1.67	130.00	0
130	L-Lysine (%)	0014	0.90500	0.09000	0.97656	0.05171	0.01674	29	-1.38	130.05	0
130	L-Lysine (%)	0859	0.90570	0.01580	0.97656	0.05171	0.01674	29	-1.37	130.00	0
130	L-Lysine (%)	0939	0.91500	0.01000	0.97656	0.05171	0.01674	29	-1.19	130.05	0
130	L-Lysine (%)	0910	0.93500	0.03000	0.97656	0.05171	0.01674	29	-0.80	130.00	0
130	L-Lysine (%)	0652	0.94000	0.00000	0.97656	0.05171	0.01674	29	-0.71	130.00	0
130	L-Lysine (%)	0878	0.96300	0.00400	0.97656	0.05171	0.01674	29	-0.26	130.00	0
130	L-Lysine (%)	0512	0.96340	0.00860	0.97656	0.05171	0.01674	29	-0.25	130.00	0
130	L-Lysine (%)	2059	0.97600	0.00000	0.97656	0.05171	0.01674	29	-0.01	130.00	0
130	L-Lysine (%)	0684	0.97750	0.00500	0.97656	0.05171	0.01674	29	0.02	130.00	0
130	L-Lysine (%)	2188	0.98000	0.00000	0.97656	0.05171	0.01674	29	0.07	130.05	0
130	L-Lysine (%)	2196	0.98200	0.00000	0.97656	0.05171	0.01674	29	0.11	130.00	0
130	L-Lysine (%)	0870	0.98725	0.02870	0.97656	0.05171	0.01674	29	0.21	130.00	0
130	L-Lysine (%)	0571	0.99300	0.00000	0.97656	0.05171	0.01674	29	0.32	130.00	0
130	L-Lysine (%)	0098	0.99850	0.00700	0.97656	0.05171	0.01674	29	0.42	130.02	0
130	L-Lysine (%)	0872	0.99900	0.00800	0.97656	0.05171	0.01674	29	0.43	130.00	0
130	L-Lysine (%)	0644	1.0005	0.00100	0.97656	0.05171	0.01674	29	0.46	130.00	0
130	L-Lysine (%)	0676	1.0050	0.01000	0.97656	0.05171	0.01674	29	0.55	130.05	0
130	L-Lysine (%)	0918	1.0068	0.06940	0.97656	0.05171	0.01674	29	0.58	130.00	0
130	L-Lysine (%)	2195	1.0100	0.00000	0.97656	0.05171	0.01674	29	0.65	130.05	0
130	L-Lysine (%)	0675	1.0150	0.01000	0.97656	0.05171	0.01674	29	0.74	130.00	0
130	L-Lysine (%)	0354	1.0180	0.00600	0.97656	0.05171	0.01674	29	0.80	130.00	0
130	L-Lysine (%)	0723	1.0210	0.00000	0.97656	0.05171	0.01674	29	0.86	130.05	0
130	L-Lysine (%)	0968	1.0225	0.00500	0.97656	0.05171	0.01674	29	0.89	130.00	0
130	L-Lysine (%)	0868	1.0230	0.00400	0.97656	0.05171	0.01674	29	0.90	130.00	0
130	L-Lysine (%)	0941	1.0270	0.04400	0.97656	0.05171	0.01674	29	0.98	130.00	0
130	L-Lysine (%)	0504	1.0950	0.01000	0.97656	0.05171	0.01674	29	2.29	130.00	0
130	L-Lysine (%)	0027	1.0845	0.22300	0.97656	0.05171	0.01674	29	2.09	130.05	1
131	Methionine (%)	2053	0.23200	0.00400	0.30215	0.02679	0.00600	27	-2.62	131.99	0
131	Methionine (%)	0868	0.23850	0.00300	0.30215	0.02679	0.00600	27	-2.38	131.00	0
131	Methionine (%)	0939	0.25000	0.02000	0.30215	0.02679	0.00600	27	-1.95	131.05	0
131	Methionine (%)	2195	0.26000	0.00000	0.30215	0.02679	0.00600	27	-1.57	131.05	0
131	Methionine (%)	0652	0.27000	0.00000	0.30215	0.02679	0.00600	27	-1.20	131.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
131	Methionine (%)	0859	0.27905	0.00490	0.30215	0.02679	0.00600	27	-0.86	131.00	0
131	Methionine (%)	0684	0.29600	0.00400	0.30215	0.02679	0.00600	27	-0.23	131.00	0
131	Methionine (%)	0014	0.29600	0.01800	0.30215	0.02679	0.00600	27	-0.23	131.05	0
131	Methionine (%)	2059	0.29650	0.00300	0.30215	0.02679	0.00600	27	-0.21	131.00	0
131	Methionine (%)	0098	0.29650	0.02300	0.30215	0.02679	0.00600	27	-0.21	131.02	0
131	Methionine (%)	0171	0.30000	0.00200	0.30215	0.02679	0.00600	27	-0.08	131.00	0
131	Methionine (%)	0504	0.30000	0.00000	0.30215	0.02679	0.00600	27	-0.08	131.00	0
131	Methionine (%)	2196	0.30100	0.00000	0.30215	0.02679	0.00600	27	-0.04	131.00	0
131	Methionine (%)	0941	0.30250	0.00500	0.30215	0.02679	0.00600	27	0.01	131.00	0
131	Methionine (%)	0878	0.30400	0.00200	0.30215	0.02679	0.00600	27	0.07	131.00	0
131	Methionine (%)	0723	0.30680	0.00000	0.30215	0.02679	0.00600	27	0.17	131.05	0
131	Methionine (%)	0968	0.30950	0.00300	0.30215	0.02679	0.00600	27	0.27	131.00	0
131	Methionine (%)	0571	0.31450	0.00300	0.30215	0.02679	0.00600	27	0.46	131.00	0
131	Methionine (%)	0354	0.31500	0.00000	0.30215	0.02679	0.00600	27	0.48	131.00	0
131	Methionine (%)	0910	0.31500	0.01000	0.30215	0.02679	0.00600	27	0.48	131.00	0
131	Methionine (%)	0872	0.31750	0.00100	0.30215	0.02679	0.00600	27	0.57	131.00	0
131	Methionine (%)	0870	0.32190	0.00460	0.30215	0.02679	0.00600	27	0.74	131.00	0
131	Methionine (%)	0918	0.32500	0.01000	0.30215	0.02679	0.00600	27	0.85	131.00	0
131	Methionine (%)	0675	0.33000	0.02000	0.30215	0.02679	0.00600	27	1.04	131.00	0
131	Methionine (%)	0512	0.33410	0.01160	0.30215	0.02679	0.00600	27	1.19	131.00	0
131	Methionine (%)	0644	0.33700	0.01000	0.30215	0.02679	0.00600	27	1.30	131.00	0
131	Methionine (%)	2188	0.37000	0.00000	0.30215	0.02679	0.00600	27	2.53	131.05	0
132	Phenylalanine (%)	0939	0.71500	0.07000	0.84226	0.03410	0.01453	23	-3.73	132.05	0
132	Phenylalanine (%)	0910	0.79500	0.07000	0.84226	0.03410	0.01453	23	-1.39	132.00	0
132	Phenylalanine (%)	0878	0.79700	0.00200	0.84226	0.03410	0.01453	23	-1.33	132.00	0
132	Phenylalanine (%)	0676	0.80000	0.00000	0.84226	0.03410	0.01453	23	-1.24	132.05	0
132	Phenylalanine (%)	0870	0.80860	0.00480	0.84226	0.03410	0.01453	23	-0.99	132.00	0
132	Phenylalanine (%)	0872	0.82700	0.00200	0.84226	0.03410	0.01453	23	-0.45	132.00	0
132	Phenylalanine (%)	0684	0.82850	0.00300	0.84226	0.03410	0.01453	23	-0.40	132.00	0
132	Phenylalanine (%)	2188	0.83000	0.02000	0.84226	0.03410	0.01453	23	-0.36	132.05	0
132	Phenylalanine (%)	2196	0.83800	0.00000	0.84226	0.03410	0.01453	23	-0.12	132.00	0
132	Phenylalanine (%)	0098	0.84150	0.01300	0.84226	0.03410	0.01453	23	-0.02	132.02	0
132	Phenylalanine (%)	0868	0.84200	0.02000	0.84226	0.03410	0.01453	23	-0.01	132.00	0
132	Phenylalanine (%)	0941	0.84500	0.04200	0.84226	0.03410	0.01453	23	0.08	132.00	0
132	Phenylalanine (%)	0354	0.84600	0.00000	0.84226	0.03410	0.01453	23	0.11	132.00	0
132	Phenylalanine (%)	0171	0.84950	0.00100	0.84226	0.03410	0.01453	23	0.21	132.00	0
132	Phenylalanine (%)	0859	0.85085	0.01690	0.84226	0.03410	0.01453	23	0.25	132.00	0
132	Phenylalanine (%)	0968	0.85300	0.00000	0.84226	0.03410	0.01453	23	0.32	132.00	0
132	Phenylalanine (%)	0644	0.85550	0.00500	0.84226	0.03410	0.01453	23	0.39	132.00	0
132	Phenylalanine (%)	0571	0.85750	0.01500	0.84226	0.03410	0.01453	23	0.45	132.00	0
132	Phenylalanine (%)	0504	0.87000	0.00000	0.84226	0.03410	0.01453	23	0.81	132.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
132	Phenylalanine (%)	0675	0.87500	0.01000	0.84226	0.03410	0.01453	23	0.96	132.00	0
132	Phenylalanine (%)	2059	0.88400	0.00000	0.84226	0.03410	0.01453	23	1.22	132.00	0
132	Phenylalanine (%)	0918	0.89665	0.01950	0.84226	0.03410	0.01453	23	1.60	132.00	0
132	Phenylalanine (%)	0652	0.91000	0.02000	0.84226	0.03410	0.01453	23	1.99	132.00	0
133	Proline (%)	0859	0.81775	0.02290	1.0476	0.07699	0.01725	22	-2.99	133.00	0
133	Proline (%)	0870	0.92045	0.02610	1.0476	0.07699	0.01725	22	-1.65	133.00	0
133	Proline (%)	0171	0.96200	0.00200	1.0476	0.07699	0.01725	22	-1.11	133.00	0
133	Proline (%)	0652	0.97500	0.03000	1.0476	0.07699	0.01725	22	-0.94	133.00	0
133	Proline (%)	0910	0.97500	0.05000	1.0476	0.07699	0.01725	22	-0.94	133.00	0
133	Proline (%)	0878	1.0010	0.01000	1.0476	0.07699	0.01725	22	-0.61	133.00	0
133	Proline (%)	0968	1.0160	0.00400	1.0476	0.07699	0.01725	22	-0.41	133.00	0
133	Proline (%)	0918	1.0296	0.02360	1.0476	0.07699	0.01725	22	-0.23	133.00	0
133	Proline (%)	2059	1.0315	0.00900	1.0476	0.07699	0.01725	22	-0.21	133.00	0
133	Proline (%)	0354	1.0445	0.00500	1.0476	0.07699	0.01725	22	-0.04	133.00	0
133	Proline (%)	0571	1.0445	0.00900	1.0476	0.07699	0.01725	22	-0.04	133.00	0
133	Proline (%)	0872	1.0560	0.01800	1.0476	0.07699	0.01725	22	0.11	133.00	0
133	Proline (%)	0644	1.0630	0.02600	1.0476	0.07699	0.01725	22	0.20	133.00	0
133	Proline (%)	0939	1.0650	0.03000	1.0476	0.07699	0.01725	22	0.23	133.05	0
133	Proline (%)	0676	1.0750	0.01000	1.0476	0.07699	0.01725	22	0.36	133.05	0
133	Proline (%)	0684	1.0765	0.01900	1.0476	0.07699	0.01725	22	0.38	133.00	0
133	Proline (%)	0504	1.0900	0.00000	1.0476	0.07699	0.01725	22	0.55	133.00	0
133	Proline (%)	0941	1.0910	0.04200	1.0476	0.07699	0.01725	22	0.56	133.00	0
133	Proline (%)	0868	1.0985	0.00300	1.0476	0.07699	0.01725	22	0.66	133.00	0
133	Proline (%)	0675	1.1650	0.03000	1.0476	0.07699	0.01725	22	1.52	133.00	0
133	Proline (%)	2196	1.2140	0.00000	1.0476	0.07699	0.01725	22	2.16	133.00	0
133	Proline (%)	2188	1.2650	0.01000	1.0476	0.07699	0.01725	22	2.82	133.05	0
134	Serine (%)	0939	0.74000	0.06000	0.84685	0.04119	0.01645	22	-2.59	134.05	0
134	Serine (%)	0859	0.79320	0.01460	0.84685	0.04119	0.01645	22	-1.30	134.00	0
134	Serine (%)	0098	0.80450	0.01700	0.84685	0.04119	0.01645	22	-1.03	134.02	0
134	Serine (%)	0870	0.80760	0.01120	0.84685	0.04119	0.01645	22	-0.95	134.00	0
134	Serine (%)	0652	0.81500	0.03000	0.84685	0.04119	0.01645	22	-0.77	134.00	0
134	Serine (%)	0968	0.82050	0.00300	0.84685	0.04119	0.01645	22	-0.64	134.00	0
134	Serine (%)	0878	0.82550	0.01500	0.84685	0.04119	0.01645	22	-0.52	134.00	0
134	Serine (%)	0644	0.82750	0.00500	0.84685	0.04119	0.01645	22	-0.47	134.00	0
134	Serine (%)	2059	0.82900	0.00400	0.84685	0.04119	0.01645	22	-0.43	134.00	0
134	Serine (%)	2196	0.84100	0.00000	0.84685	0.04119	0.01645	22	-0.14	134.00	0
134	Serine (%)	0571	0.84600	0.00200	0.84685	0.04119	0.01645	22	-0.02	134.00	0
134	Serine (%)	0675	0.85000	0.02000	0.84685	0.04119	0.01645	22	0.08	134.00	0
134	Serine (%)	2188	0.85000	0.02000	0.84685	0.04119	0.01645	22	0.08	134.05	0
134	Serine (%)	0872	0.85500	0.00800	0.84685	0.04119	0.01645	22	0.20	134.00	0
134	Serine (%)	0910	0.86000	0.04000	0.84685	0.04119	0.01645	22	0.32	134.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
134	Serine (%)	0941	0.86350	0.01500	0.84685	0.04119	0.01645	22	0.40	134.00	0
134	Serine (%)	0354	0.87400	0.00200	0.84685	0.04119	0.01645	22	0.66	134.00	0
134	Serine (%)	0868	0.87550	0.01100	0.84685	0.04119	0.01645	22	0.70	134.00	0
134	Serine (%)	0684	0.88200	0.03200	0.84685	0.04119	0.01645	22	0.85	134.00	0
134	Serine (%)	0171	0.90950	0.00500	0.84685	0.04119	0.01645	22	1.52	134.00	0
134	Serine (%)	0918	0.91060	0.00700	0.84685	0.04119	0.01645	22	1.55	134.00	0
134	Serine (%)	0676	0.92000	0.04000	0.84685	0.04119	0.01645	22	1.78	134.05	0
134	Serine (%)	0504	0.73000	0.10000	0.84685	0.04119	0.01645	22	-2.84	134.00	1
135	Threonine (%)	0859	0.62690	0.01340	0.69177	0.03151	0.00580	23	-2.06	135.00	0
135	Threonine (%)	0652	0.64000	0.00000	0.69177	0.03151	0.00580	23	-1.64	135.00	0
135	Threonine (%)	2195	0.66000	0.00000	0.69177	0.03151	0.00580	23	-1.01	135.05	0
135	Threonine (%)	0878	0.66050	0.00700	0.69177	0.03151	0.00580	23	-0.99	135.00	0
135	Threonine (%)	0910	0.66500	0.01000	0.69177	0.03151	0.00580	23	-0.85	135.00	0
135	Threonine (%)	0968	0.67000	0.00000	0.69177	0.03151	0.00580	23	-0.69	135.00	0
135	Threonine (%)	0870	0.67905	0.00250	0.69177	0.03151	0.00580	23	-0.40	135.00	0
135	Threonine (%)	0872	0.67950	0.00300	0.69177	0.03151	0.00580	23	-0.39	135.00	0
135	Threonine (%)	2059	0.68250	0.00300	0.69177	0.03151	0.00580	23	-0.29	135.00	0
135	Threonine (%)	0644	0.68300	0.00400	0.69177	0.03151	0.00580	23	-0.28	135.00	0
135	Threonine (%)	0571	0.68400	0.00600	0.69177	0.03151	0.00580	23	-0.25	135.00	0
135	Threonine (%)	0098	0.69100	0.00600	0.69177	0.03151	0.00580	23	-0.02	135.02	0
135	Threonine (%)	2196	0.69200	0.00000	0.69177	0.03151	0.00580	23	0.01	135.00	0
135	Threonine (%)	0941	0.69650	0.01300	0.69177	0.03151	0.00580	23	0.15	135.00	0
135	Threonine (%)	0939	0.70000	0.00000	0.69177	0.03151	0.00580	23	0.26	135.05	0
135	Threonine (%)	0868	0.70450	0.00500	0.69177	0.03151	0.00580	23	0.40	135.00	0
135	Threonine (%)	0354	0.70650	0.00300	0.69177	0.03151	0.00580	23	0.47	135.00	0
135	Threonine (%)	0675	0.71500	0.01000	0.69177	0.03151	0.00580	23	0.74	135.00	0
135	Threonine (%)	0684	0.72100	0.00400	0.69177	0.03151	0.00580	23	0.93	135.00	0
135	Threonine (%)	0918	0.72360	0.01360	0.69177	0.03151	0.00580	23	1.01	135.00	0
135	Threonine (%)	0676	0.73000	0.02000	0.69177	0.03151	0.00580	23	1.21	135.05	0
135	Threonine (%)	0171	0.75000	0.00000	0.69177	0.03151	0.00580	23	1.85	135.00	0
135	Threonine (%)	2188	0.75500	0.01000	0.69177	0.03151	0.00580	23	2.01	135.05	0
135	Threonine (%)	0504	0.66500	0.03000	0.69177	0.03151	0.00580	23	-0.85	135.00	1
136	Tryptophan (%)	0870	0.18020	0.00060	0.25171	0.02603	0.00493	13	-2.75	136.00	0
136	Tryptophan (%)	0910	0.22000	0.00000	0.25171	0.02603	0.00493	13	-1.22	136.00	0
136	Tryptophan (%)	0504	0.22000	0.00000	0.25171	0.02603	0.00493	13	-1.22	136.99	0
136	Tryptophan (%)	0868	0.23200	0.01000	0.25171	0.02603	0.00493	13	-0.76	136.01	0
136	Tryptophan (%)	0098	0.24650	0.01100	0.25171	0.02603	0.00493	13	-0.20	136.02	0
136	Tryptophan (%)	0918	0.26000	0.02000	0.25171	0.02603	0.00493	13	0.32	136.00	0
136	Tryptophan (%)	2059	0.26300	0.00200	0.25171	0.02603	0.00493	13	0.43	136.03	0
136	Tryptophan (%)	0878	0.26400	0.00200	0.25171	0.02603	0.00493	13	0.47	136.01	0
136	Tryptophan (%)	0872	0.26450	0.00500	0.25171	0.02603	0.00493	13	0.49	136.03	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
136	Tryptophan (%)	0859	0.26655	0.00550	0.25171	0.02603	0.00493	13	0.57	136.03	0
136	Tryptophan (%)	0684	0.26700	0.00600	0.25171	0.02603	0.00493	13	0.59	136.00	0
136	Tryptophan (%)	0571	0.27450	0.00100	0.25171	0.02603	0.00493	13	0.88	136.03	0
136	Tryptophan (%)	0644	0.28150	0.00100	0.25171	0.02603	0.00493	13	1.14	136.01	0
137	Tyrosine (%)	0939	0.42000	0.00000	0.54279	0.06492	0.01271	15	-1.89	137.05	0
137	Tyrosine (%)	0870	0.45700	0.00820	0.54279	0.06492	0.01271	15	-1.32	137.02	0
137	Tyrosine (%)	0910	0.47500	0.05000	0.54279	0.06492	0.01271	15	-1.04	137.00	0
137	Tyrosine (%)	0968	0.50000	0.00000	0.54279	0.06492	0.01271	15	-0.66	137.99	0
137	Tyrosine (%)	0171	0.50400	0.00200	0.54279	0.06492	0.01271	15	-0.60	137.99	0
137	Tyrosine (%)	0354	0.51000	0.00000	0.54279	0.06492	0.01271	15	-0.51	137.02	0
137	Tyrosine (%)	0872	0.55250	0.00500	0.54279	0.06492	0.01271	15	0.15	137.00	0
137	Tyrosine (%)	0504	0.56500	0.01000	0.54279	0.06492	0.01271	15	0.34	137.99	0
137	Tyrosine (%)	0918	0.57140	0.03440	0.54279	0.06492	0.01271	15	0.44	137.02	0
137	Tyrosine (%)	0098	0.57300	0.01200	0.54279	0.06492	0.01271	15	0.47	137.02	0
137	Tyrosine (%)	2059	0.57500	0.00600	0.54279	0.06492	0.01271	15	0.50	137.00	0
137	Tyrosine (%)	0644	0.57550	0.02300	0.54279	0.06492	0.01271	15	0.50	137.00	0
137	Tyrosine (%)	2196	0.58800	0.00000	0.54279	0.06492	0.01271	15	0.70	137.02	0
137	Tyrosine (%)	0676	0.62000	0.02000	0.54279	0.06492	0.01271	15	1.19	137.05	0
137	Tyrosine (%)	2188	0.63000	0.02000	0.54279	0.06492	0.01271	15	1.34	137.05	0
138	Valine (%)	0870	0.78520	0.02300	0.88720	0.03499	0.01507	23	-2.92	138.00	0
138	Valine (%)	0652	0.81500	0.05000	0.88720	0.03499	0.01507	23	-2.06	138.00	0
138	Valine (%)	0939	0.85000	0.02000	0.88720	0.03499	0.01507	23	-1.06	138.05	0
138	Valine (%)	0878	0.86000	0.01200	0.88720	0.03499	0.01507	23	-0.78	138.00	0
138	Valine (%)	2059	0.86300	0.00800	0.88720	0.03499	0.01507	23	-0.69	138.00	0
138	Valine (%)	0872	0.86350	0.00500	0.88720	0.03499	0.01507	23	-0.68	138.00	0
138	Valine (%)	0859	0.86430	0.01740	0.88720	0.03499	0.01507	23	-0.65	138.00	0
138	Valine (%)	0868	0.86600	0.01400	0.88720	0.03499	0.01507	23	-0.61	138.00	0
138	Valine (%)	0684	0.86900	0.01200	0.88720	0.03499	0.01507	23	-0.52	138.00	0
138	Valine (%)	2188	0.87000	0.02000	0.88720	0.03499	0.01507	23	-0.49	138.05	0
138	Valine (%)	0910	0.89000	0.04000	0.88720	0.03499	0.01507	23	0.08	138.00	0
138	Valine (%)	0171	0.89150	0.00100	0.88720	0.03499	0.01507	23	0.12	138.00	0
138	Valine (%)	0571	0.89400	0.01200	0.88720	0.03499	0.01507	23	0.19	138.00	0
138	Valine (%)	0354	0.89500	0.00000	0.88720	0.03499	0.01507	23	0.22	138.00	0
138	Valine (%)	0918	0.89605	0.02530	0.88720	0.03499	0.01507	23	0.25	138.00	0
138	Valine (%)	2196	0.90200	0.00000	0.88720	0.03499	0.01507	23	0.42	138.00	0
138	Valine (%)	0968	0.90500	0.01000	0.88720	0.03499	0.01507	23	0.51	138.00	0
138	Valine (%)	0941	0.91450	0.02900	0.88720	0.03499	0.01507	23	0.78	138.00	0
138	Valine (%)	0098	0.91500	0.01800	0.88720	0.03499	0.01507	23	0.79	138.02	0
138	Valine (%)	0644	0.91800	0.00000	0.88720	0.03499	0.01507	23	0.88	138.00	0
138	Valine (%)	0504	0.93000	0.02000	0.88720	0.03499	0.01507	23	1.22	138.00	0
138	Valine (%)	0676	0.94000	0.00000	0.88720	0.03499	0.01507	23	1.51	138.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
138	Valine (%)	0675	0.96500	0.01000	0.88720	0.03499	0.01507	23	2.22	138.00	0
139	Taurine (%)	0504	0.16000	0.00000				1		139.00	0
139	Taurine (%)	0098	0.00000	0.00000				1		139.02	4
351	Chlortetracycline (mg/kg (ppm))	0218	3.3753	0.06690				2		351.00	0
351	Chlortetracycline (mg/kg (ppm))	2127	7.6000	1.4000				2		351.05	0
354	Decoquinatate (mg/kg (ppm))	0610	0.13000	0.00000				1		354.04	0
361	Lasalocid sodium (mg/kg (ppm))	0027	0.45100	0.02200	0.61686	0.13919	0.01523	4		361.03	0
361	Lasalocid sodium (mg/kg (ppm))	2135	0.56970	0.00160	0.61686	0.13919	0.01523	4		361.05	0
361	Lasalocid sodium (mg/kg (ppm))	0610	0.67000	0.02000	0.61686	0.13919	0.01523	4		361.05	0
361	Lasalocid sodium (mg/kg (ppm))	0964	0.77675	0.01730	0.61686	0.13919	0.01523	4		361.03	0
361	Lasalocid sodium (mg/kg (ppm))	2127	98.084	195.83	0.61686	0.13919	0.01523	4		361.05	2
361	Lasalocid sodium (mg/kg (ppm))	0875	0.00000	0.00000	0.61686	0.13919	0.01523	4		361.05	4
365	Monensin (mg/kg (ppm))	0668	0.43500	0.01000	0.72327	0.29802	0.10820	7	-0.97	365.99	0
365	Monensin (mg/kg (ppm))	0668	0.45000	0.10000	0.72327	0.29802	0.10820	7	-0.92	365.02	0
365	Monensin (mg/kg (ppm))	0610	0.61500	0.01000	0.72327	0.29802	0.10820	7	-0.36	365.05	0
365	Monensin (mg/kg (ppm))	2135	0.67715	0.04090	0.72327	0.29802	0.10820	7	-0.15	365.05	0
365	Monensin (mg/kg (ppm))	2053	0.79950	0.21500	0.72327	0.29802	0.10820	7	0.26	365.04	0
365	Monensin (mg/kg (ppm))	0964	0.91595	0.38150	0.72327	0.29802	0.10820	7	0.65	365.03	0
365	Monensin (mg/kg (ppm))	0905	2.0000	0.00000	0.72327	0.29802	0.10820	7	4.28	365.02	0
365	Monensin (mg/kg (ppm))	2127	392.32	783.36	0.72327	0.29802	0.10820	7	1313.98	365.05	2
365	Monensin (mg/kg (ppm))	0875	0.00000	0.00000	0.72327	0.29802	0.10820	7		365.05	4
377	Pyrantel Tartrate (mg/kg (ppm))	2053	212.00	6.0000				1		377.02	0
392	Fenbendazole (mg/kg (ppm))	2053	96.850	3.3000				1		392.01	0
400	Water activity (Units)	2073	0.35850	0.00100	0.39746	0.04369	0.00320	7	-0.89	400.99	0
400	Water activity (Units)	0942	0.35900	0.00400	0.39746	0.04369	0.00320	7	-0.88	400.01	0
400	Water activity (Units)	0407	0.36800	0.00320	0.39746	0.04369	0.00320	7	-0.67	400.01	0
400	Water activity (Units)	0589	0.38700	0.00600	0.39746	0.04369	0.00320	7	-0.24	400.01	0
400	Water activity (Units)	2181	0.42170	0.00660	0.39746	0.04369	0.00320	7	0.55	400.01	0
400	Water activity (Units)	2109	0.44150	0.00060	0.39746	0.04369	0.00320	7	1.01	400.01	0
400	Water activity (Units)	2009	0.44650	0.00100	0.39746	0.04369	0.00320	7	1.12	400.01	0
412	Dietary Starch (%)	2129	15.670	0.36000				2		412.01	0
412	Dietary Starch (%)	0353	16.075	0.31000				2		412.01	0
516	Arsenic, total (mg / kg (ppm))	0045	0.39500	0.03000	0.46511	0.07353	0.04595	6	-0.95	516.00	0
516	Arsenic, total (mg / kg (ppm))	0171	0.40000	0.02000	0.46511	0.07353	0.04595	6	-0.89	516.00	0
516	Arsenic, total (mg / kg (ppm))	0910	0.45500	0.05000	0.46511	0.07353	0.04595	6	-0.14	516.52	0
516	Arsenic, total (mg / kg (ppm))	0186	0.45950	0.02900	0.46511	0.07353	0.04595	6	-0.08	516.52	0
516	Arsenic, total (mg / kg (ppm))	0098	0.52750	0.11300	0.46511	0.07353	0.04595	6	0.85	516.53	0
516	Arsenic, total (mg / kg (ppm))	0918	0.55365	0.03370	0.46511	0.07353	0.04595	6	1.20	516.53	0
518	Cadmium (mg / kg (ppm))	0918	0.20800	0.02980	0.22254	0.00929	0.03742	6	-1.57	518.53	0
518	Cadmium (mg / kg (ppm))	0186	0.21750	0.02500	0.22254	0.00929	0.03742	6	-0.54	518.52	0
518	Cadmium (mg / kg (ppm))	0098	0.22450	0.01500	0.22254	0.00929	0.03742	6	0.21	518.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
518	Cadmium (mg / kg (ppm))	0171	0.22500	0.01000	0.22254	0.00929	0.03742	6	0.26	518.41	0
518	Cadmium (mg / kg (ppm))	0508	0.22965	0.12470	0.22254	0.00929	0.03742	6	0.76	518.43	0
518	Cadmium (mg / kg (ppm))	0910	0.23000	0.02000	0.22254	0.00929	0.03742	6	0.80	518.52	0
520	Chromium (mg / kg (ppm))	0186	4.8600	0.12000	9.4947	3.1318	0.49488	10	-1.48	520.52	0
520	Chromium (mg / kg (ppm))	0918	5.7331	0.16680	9.4947	3.1318	0.49488	10	-1.20	520.53	0
520	Chromium (mg / kg (ppm))	0171	6.8850	0.11000	9.4947	3.1318	0.49488	10	-0.83	520.41	0
520	Chromium (mg / kg (ppm))	0693	8.8110	0.54800	9.4947	3.1318	0.49488	10	-0.22	520.42	0
520	Chromium (mg / kg (ppm))	0045	10.265	1.2700	9.4947	3.1318	0.49488	10	0.25	520.42	0
520	Chromium (mg / kg (ppm))	0955	11.100	1.0000	9.4947	3.1318	0.49488	10	0.51	520.43	0
520	Chromium (mg / kg (ppm))	0508	11.393	0.55400	9.4947	3.1318	0.49488	10	0.61	520.43	0
520	Chromium (mg / kg (ppm))	2053	11.490	0.92000	9.4947	3.1318	0.49488	10	0.64	520.42	0
520	Chromium (mg / kg (ppm))	0510	11.825	0.05000	9.4947	3.1318	0.49488	10	0.74	520.43	0
520	Chromium (mg / kg (ppm))	0032	12.585	0.21000	9.4947	3.1318	0.49488	10	0.99	520.53	0
526	Lead (mg / kg (ppm))	0918	0.33765	0.00290	0.37329	0.03082	0.03948	4		526.53	0
526	Lead (mg / kg (ppm))	0098	0.36200	0.03600	0.37329	0.03082	0.03948	4		526.53	0
526	Lead (mg / kg (ppm))	0186	0.38350	0.03900	0.37329	0.03082	0.03948	4		526.52	0
526	Lead (mg / kg (ppm))	0910	0.41000	0.08000	0.37329	0.03082	0.03948	4		526.52	0
526	Lead (mg / kg (ppm))	0171	0.71000	0.42000	0.37329	0.03082	0.03948	4		526.41	2
539	Nickel (mg / kg (ppm))	0186	4.4100	0.74000	5.5818	1.2020	0.48220	5		539.52	0
539	Nickel (mg / kg (ppm))	0918	4.4736	0.03990	5.5818	1.2020	0.48220	5		539.53	0
539	Nickel (mg / kg (ppm))	0171	5.5000	0.00000	5.5818	1.2020	0.48220	5		539.41	0
539	Nickel (mg / kg (ppm))	0508	6.3255	0.83110	5.5818	1.2020	0.48220	5		539.43	0
539	Nickel (mg / kg (ppm))	2053	7.2000	0.80000	5.5818	1.2020	0.48220	5		539.42	0
710	Lauric Acid (12:0) (%) (w/w)	0226	0.00285	0.00050				2		710.99	0
710	Lauric Acid (12:0) (%) (w/w)	0676	0.00550	0.00100				2		710.99	0
714	Myristic Acid (14:0) (%) (w/w)	0226	0.00940	0.00060				1		714.99	0
716	Palmitic Acid (16:0) (%) (w/w)	0226	0.64735	0.05650				1		716.99	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0676	0.00700	0.00000				2		718.99	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0226	0.06385	0.00470				2		718.99	0
722	Stearic Acid (18:0) (%) (w/w)	0226	0.14070	0.01340				1		722.99	0
724	Oleic Acid (9c-18:1) (%) (w/w)	0226	0.78495	0.07970				1		724.99	0
726	Linoleic Acid (9c,12c-18:2) (%) (w/w)	0226	0.51075	0.12350				3		726.99	0
726	Linoleic Acid (9c,12c-18:2) (%) (w/w)	0676	0.56050	0.00700				3		726.99	0
726	Linoleic Acid (9c,12c-18:2) (%) (w/w)	0610	0.71000	0.02000				3		726.99	0
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (%) (w/w)	0226	0.20805	0.02630				2		728.99	0
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (%) (w/w)	0676	0.25050	0.00300				2		728.99	0
730	Arachidic Acid (20:0) (%) (w/w)	0226	0.01335	0.00130				1		730.99	0
732	Gondoic Acid (11c-20:1) (%) (w/w)	0226	0.02200	0.00540				1		732.99	0
736	Arachidonic Acid (5c,8c,11c,14c-20:4) (%) (w/w)	0676	0.00650	0.00300				1		736.99	0
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c-20:5) (%) (w/w)	0676	0.00200	0.00000				1		740.99	0
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c-20:5) (%) (w/w)	0226	0.00000	0.00000				1		740.99	4

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
742	Behenic Acid (22:0) (% (w/w))	0226	0.01890	0.00320				1	742.99	0	
744	Erucic Acid (13c-22:1) (% (w/w))	0226	0.00110	0.00060				1	744.99	0	
746	Docosapentaenoic Acid n-3 DPA (DHA);	0226	0.00000	0.00000				0	746.99	4	
746	Docosapentaenoic Acid n-3 DPA (DHA);	0676	0.00000	0.00000				0	746.99	4	
748	Lignoceric Acid (24:0) (% (w/w))	0226	0.01210	0.00060				1	748.99	0	
750	Docosahexaenoic Acid DHA (4c,7c,10c,	0226	0.00000	0.00000				0	750.99	4	
750	Docosahexaenoic Acid DHA (4c,7c,10c,	0676	0.00000	0.00000				0	750.99	4	
752	Nervonic Acid (24:1) isomers (% (w/w))	0226	0.01885	0.00530				1	752.99	0	
754	Total n-3 Polyunsaturated (Omega-3) F _a	0676	0.25250	0.00300				3	754.99	0	
754	Total n-3 Polyunsaturated (Omega-3) F _a	0610	0.30000	0.00000				3	754.99	0	
754	Total n-3 Polyunsaturated (Omega-3) F _a	0675	0.34000	0.02000				3	754.02	0	
756	Total n-6 Polyunsaturated (Omega-6) F _a	0676	0.57200	0.00400				3	756.99	0	
756	Total n-6 Polyunsaturated (Omega-6) F _a	0610	0.70000	0.00000				3	756.99	0	
756	Total n-6 Polyunsaturated (Omega-6) F _a	0675	1.0950	0.15000				3	756.02	0	
772	Total Fatty Acids (% (w/w))	0226	2.5254	0.27740				1	772.99	0	

Notes: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2 and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = Rejected for duplicates too far apart, 2 = Rejected as outlier, 5 = A reporting limit and 4 = zeros submitted as values. Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). Flag 3 indicates not used in statistics.