

Pig Feed, Medicated

Animal Feed Scheme

Labs Reporting: 194

Test Material Code # 201823

Analyte Proficiency Testing Report

Issue Date : 04/30/2018

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
000	Urea (%)	0278	0.15000	0.10000				1		000.02	0
001	Loss on Drying (%)	0618	6.3800	0.14000	8.2282	0.27623	0.11028	73	-6.69	001.99	0
001	Loss on Drying (%)	0536	6.8900	0.00000	8.2282	0.27623	0.11028	73	-4.84	001.99	0
001	Loss on Drying (%)	2155	7.1950	0.31000	8.2282	0.27623	0.11028	73	-3.74	001.99	0
001	Loss on Drying (%)	0038	7.2750	0.87000	8.2282	0.27623	0.11028	73	-3.45	001.07	0
001	Loss on Drying (%)	0874	7.4800	0.12000	8.2282	0.27623	0.11028	73	-2.71	001.02	0
001	Loss on Drying (%)	2195	7.6200	0.04000	8.2282	0.27623	0.11028	73	-2.20	001.07	0
001	Loss on Drying (%)	0676	7.6350	0.03000	8.2282	0.27623	0.11028	73	-2.15	001.99	0
001	Loss on Drying (%)	0610	7.7200	0.06000	8.2282	0.27623	0.11028	73	-1.84	001.05	0
001	Loss on Drying (%)	0889	7.7700	0.02000	8.2282	0.27623	0.11028	73	-1.66	001.99	0
001	Loss on Drying (%)	0074	7.8550	0.17000	8.2282	0.27623	0.11028	73	-1.35	001.07	0
001	Loss on Drying (%)	0630	7.8550	0.15000	8.2282	0.27623	0.11028	73	-1.35	001.99	0
001	Loss on Drying (%)	2146	7.9850	0.19000	8.2282	0.27623	0.11028	73	-0.88	001.07	0
001	Loss on Drying (%)	0510	8.0000	0.20000	8.2282	0.27623	0.11028	73	-0.83	001.99	0
001	Loss on Drying (%)	2145	8.0150	0.01000	8.2282	0.27623	0.11028	73	-0.77	001.07	0
001	Loss on Drying (%)	0918	8.0150	0.15000	8.2282	0.27623	0.11028	73	-0.77	001.99	0
001	Loss on Drying (%)	0675	8.0400	0.08000	8.2282	0.27623	0.11028	73	-0.68	001.07	0
001	Loss on Drying (%)	0940	8.0450	0.09000	8.2282	0.27623	0.11028	73	-0.66	001.99	0
001	Loss on Drying (%)	0912	8.0900	0.08000	8.2282	0.27623	0.11028	73	-0.50	001.99	0
001	Loss on Drying (%)	0353	8.1000	0.36000	8.2282	0.27623	0.11028	73	-0.46	001.07	0
001	Loss on Drying (%)	0278	8.1050	0.15000	8.2282	0.27623	0.11028	73	-0.45	001.07	0
001	Loss on Drying (%)	0407	8.1057	0.01280	8.2282	0.27623	0.11028	73	-0.44	001.07	0
001	Loss on Drying (%)	0083	8.1100	0.10000	8.2282	0.27623	0.11028	73	-0.43	001.07	0
001	Loss on Drying (%)	2109	8.1200	0.30000	8.2282	0.27623	0.11028	73	-0.39	001.07	0
001	Loss on Drying (%)	0948	8.1350	0.05000	8.2282	0.27623	0.11028	73	-0.34	001.99	0
001	Loss on Drying (%)	0226	8.1400	0.40000	8.2282	0.27623	0.11028	73	-0.32	001.07	0
001	Loss on Drying (%)	2190	8.1550	0.01000	8.2282	0.27623	0.11028	73	-0.27	001.07	0
001	Loss on Drying (%)	2126	8.1600	0.02000	8.2282	0.27623	0.11028	73	-0.25	001.99	0
001	Loss on Drying (%)	0186	8.1750	0.11000	8.2282	0.27623	0.11028	73	-0.19	001.03	0
001	Loss on Drying (%)	0297	8.1800	0.04000	8.2282	0.27623	0.11028	73	-0.17	001.07	0
001	Loss on Drying (%)	2059	8.1950	0.09000	8.2282	0.27623	0.11028	73	-0.12	001.07	0
001	Loss on Drying (%)	0413	8.2000	0.20000	8.2282	0.27623	0.11028	73	-0.10	001.07	0
001	Loss on Drying (%)	0609	8.2000	0.00000	8.2282	0.27623	0.11028	73	-0.10	001.07	0
001	Loss on Drying (%)	0505	8.2050	0.05000	8.2282	0.27623	0.11028	73	-0.08	001.07	0
001	Loss on Drying (%)	0683	8.2050	0.15000	8.2282	0.27623	0.11028	73	-0.08	001.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
001	Loss on Drying (%)	0590	8.2100	0.04000	8.2282	0.27623	0.11028	73	-0.07	001.07	0
001	Loss on Drying (%)	0171	8.2150	0.03000	8.2282	0.27623	0.11028	73	-0.05	001.07	0
001	Loss on Drying (%)	0003	8.2200	0.04000	8.2282	0.27623	0.11028	73	-0.03	001.07	0
001	Loss on Drying (%)	0504	8.2350	0.03000	8.2282	0.27623	0.11028	73	0.02	001.00	0
001	Loss on Drying (%)	0720	8.2350	0.09000	8.2282	0.27623	0.11028	73	0.02	001.99	0
001	Loss on Drying (%)	0139	8.2400	0.04000	8.2282	0.27623	0.11028	73	0.04	001.07	0
001	Loss on Drying (%)	0629	8.2500	0.26000	8.2282	0.27623	0.11028	73	0.08	001.99	0
001	Loss on Drying (%)	0895	8.2550	0.01000	8.2282	0.27623	0.11028	73	0.10	001.03	0
001	Loss on Drying (%)	0034	8.2550	0.05000	8.2282	0.27623	0.11028	73	0.10	001.99	0
001	Loss on Drying (%)	0049	8.2700	0.12000	8.2282	0.27623	0.11028	73	0.15	001.07	0
001	Loss on Drying (%)	2192	8.3250	0.21000	8.2282	0.27623	0.11028	73	0.35	001.07	0
001	Loss on Drying (%)	0027	8.3265	0.02900	8.2282	0.27623	0.11028	73	0.36	001.00	0
001	Loss on Drying (%)	2009	8.3291	0.01910	8.2282	0.27623	0.11028	73	0.36	001.07	0
001	Loss on Drying (%)	0035	8.3300	0.04000	8.2282	0.27623	0.11028	73	0.37	001.07	0
001	Loss on Drying (%)	0571	8.3350	0.05000	8.2282	0.27623	0.11028	73	0.39	001.07	0
001	Loss on Drying (%)	0689	8.3500	0.10000	8.2282	0.27623	0.11028	73	0.44	001.07	0
001	Loss on Drying (%)	0910	8.3500	0.10000	8.2282	0.27623	0.11028	73	0.44	001.99	0
001	Loss on Drying (%)	0868	8.3650	0.01000	8.2282	0.27623	0.11028	73	0.50	001.03	0
001	Loss on Drying (%)	0015	8.3650	0.09000	8.2282	0.27623	0.11028	73	0.50	001.07	0
001	Loss on Drying (%)	0169	8.3700	0.00000	8.2282	0.27623	0.11028	73	0.51	001.00	0
001	Loss on Drying (%)	2053	8.4000	0.06000	8.2282	0.27623	0.11028	73	0.62	001.07	0
001	Loss on Drying (%)	0187	8.4050	0.01000	8.2282	0.27623	0.11028	73	0.64	001.07	0
001	Loss on Drying (%)	0644	8.4450	0.03000	8.2282	0.27623	0.11028	73	0.78	001.05	0
001	Loss on Drying (%)	0852	8.4550	0.01000	8.2282	0.27623	0.11028	73	0.82	001.99	0
001	Loss on Drying (%)	0581	8.4650	0.09000	8.2282	0.27623	0.11028	73	0.86	001.07	0
001	Loss on Drying (%)	0843	8.4650	0.29000	8.2282	0.27623	0.11028	73	0.86	001.07	0
001	Loss on Drying (%)	0849	8.4750	0.09000	8.2282	0.27623	0.11028	73	0.89	001.07	0
001	Loss on Drying (%)	0878	8.4850	0.03000	8.2282	0.27623	0.11028	73	0.93	001.03	0
001	Loss on Drying (%)	0089	8.4900	0.00000	8.2282	0.27623	0.11028	73	0.95	001.07	0
001	Loss on Drying (%)	0309	8.5000	0.06000	8.2282	0.27623	0.11028	73	0.98	001.00	0
001	Loss on Drying (%)	0037	8.5050	0.01000	8.2282	0.27623	0.11028	73	1.00	001.99	0
001	Loss on Drying (%)	0142	8.5300	0.00000	8.2282	0.27623	0.11028	73	1.09	001.07	0
001	Loss on Drying (%)	0019	8.6400	0.06000	8.2282	0.27623	0.11028	73	1.49	001.07	0
001	Loss on Drying (%)	0693	8.6500	0.10000	8.2282	0.27623	0.11028	73	1.53	001.07	0
001	Loss on Drying (%)	0872	8.6800	0.04000	8.2282	0.27623	0.11028	73	1.64	001.07	0
001	Loss on Drying (%)	0914	8.8500	0.26000	8.2282	0.27623	0.11028	73	2.25	001.99	0
001	Loss on Drying (%)	0045	9.0150	0.57000	8.2282	0.27623	0.11028	73	2.85	001.07	0
001	Loss on Drying (%)	0357	9.0462	0.10990	8.2282	0.27623	0.11028	73	2.96	001.99	0
001	Loss on Drying (%)	0596	9.1600	0.12000	8.2282	0.27623	0.11028	73	3.37	001.00	0
001	Loss on Drying (%)	0560	7.3400	1.9800	8.2282	0.27623	0.11028	73	-3.22	001.99	1
001	Loss on Drying (%)	0618	18.688	0.10600	8.2282	0.27623	0.11028	73	37.87	001.07	2
002	Protein, Crude (%)	0852	15.715	0.09000	18.610	0.31768	0.19568	195	-9.11	002.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein, Crude (%)	0202	16.695	2.9300	18.610	0.31768	0.19568	195	-6.03	002.06	0
002	Protein, Crude (%)	0889	16.740	0.04000	18.610	0.31768	0.19568	195	-5.89	002.11	0
002	Protein, Crude (%)	2128	17.316	0.93200	18.610	0.31768	0.19568	195	-4.07	002.00	0
002	Protein, Crude (%)	0169	17.555	0.05000	18.610	0.31768	0.19568	195	-3.32	002.02	0
002	Protein, Crude (%)	0874	17.560	0.16000	18.610	0.31768	0.19568	195	-3.31	002.04	0
002	Protein, Crude (%)	0952	17.590	0.32000	18.610	0.31768	0.19568	195	-3.21	002.05	0
002	Protein, Crude (%)	0622	17.693	0.36270	18.610	0.31768	0.19568	195	-2.89	002.05	0
002	Protein, Crude (%)	0676	17.910	0.20000	18.610	0.31768	0.19568	195	-2.20	002.06	0
002	Protein, Crude (%)	0581	17.995	0.31000	18.610	0.31768	0.19568	195	-1.94	002.06	0
002	Protein, Crude (%)	2155	18.003	0.12400	18.610	0.31768	0.19568	195	-1.91	002.06	0
002	Protein, Crude (%)	0651	18.014	0.04700	18.610	0.31768	0.19568	195	-1.88	002.05	0
002	Protein, Crude (%)	2109	18.085	0.05000	18.610	0.31768	0.19568	195	-1.65	002.05	0
002	Protein, Crude (%)	0536	18.130	0.22000	18.610	0.31768	0.19568	195	-1.51	002.05	0
002	Protein, Crude (%)	2178	18.142	0.07800	18.610	0.31768	0.19568	195	-1.47	002.06	0
002	Protein, Crude (%)	0885	18.144	0.01690	18.610	0.31768	0.19568	195	-1.47	002.05	0
002	Protein, Crude (%)	2188	18.155	0.09000	18.610	0.31768	0.19568	195	-1.43	002.01	0
002	Protein, Crude (%)	0629	18.195	0.01000	18.610	0.31768	0.19568	195	-1.31	002.10	0
002	Protein, Crude (%)	0956	18.200	0.40000	18.610	0.31768	0.19568	195	-1.29	002.06	0
002	Protein, Crude (%)	2192	18.230	0.08000	18.610	0.31768	0.19568	195	-1.20	002.05	0
002	Protein, Crude (%)	0539	18.240	0.44000	18.610	0.31768	0.19568	195	-1.17	002.06	0
002	Protein, Crude (%)	0098	18.245	0.27000	18.610	0.31768	0.19568	195	-1.15	002.08	0
002	Protein, Crude (%)	0043	18.247	0.20700	18.610	0.31768	0.19568	195	-1.15	002.06	0
002	Protein, Crude (%)	0413	18.250	0.10000	18.610	0.31768	0.19568	195	-1.13	002.06	0
002	Protein, Crude (%)	2190	18.285	0.19000	18.610	0.31768	0.19568	195	-1.02	002.05	0
002	Protein, Crude (%)	0043	18.290	0.32000	18.610	0.31768	0.19568	195	-1.01	002.01	0
002	Protein, Crude (%)	0685	18.300	0.02000	18.610	0.31768	0.19568	195	-0.98	002.01	0
002	Protein, Crude (%)	2146	18.300	0.20000	18.610	0.31768	0.19568	195	-0.98	002.05	0
002	Protein, Crude (%)	0187	18.310	0.06000	18.610	0.31768	0.19568	195	-0.95	002.04	0
002	Protein, Crude (%)	0100	18.320	0.08000	18.610	0.31768	0.19568	195	-0.91	002.06	0
002	Protein, Crude (%)	0895	18.325	0.01000	18.610	0.31768	0.19568	195	-0.90	002.05	0
002	Protein, Crude (%)	0596	18.335	0.09000	18.610	0.31768	0.19568	195	-0.87	002.05	0
002	Protein, Crude (%)	0164	18.340	0.18000	18.610	0.31768	0.19568	195	-0.85	002.01	0
002	Protein, Crude (%)	2181	18.345	0.01000	18.610	0.31768	0.19568	195	-0.84	002.05	0
002	Protein, Crude (%)	0910	18.350	0.10000	18.610	0.31768	0.19568	195	-0.82	002.01	0
002	Protein, Crude (%)	0171	18.350	0.02000	18.610	0.31768	0.19568	195	-0.82	002.06	0
002	Protein, Crude (%)	0226	18.350	0.10000	18.610	0.31768	0.19568	195	-0.82	002.06	0
002	Protein, Crude (%)	0653	18.360	0.06000	18.610	0.31768	0.19568	195	-0.79	002.01	0
002	Protein, Crude (%)	0018	18.360	0.16000	18.610	0.31768	0.19568	195	-0.79	002.06	0
002	Protein, Crude (%)	0083	18.360	0.30000	18.610	0.31768	0.19568	195	-0.79	002.06	0
002	Protein, Crude (%)	2009	18.360	0.11170	18.610	0.31768	0.19568	195	-0.79	002.05	0
002	Protein, Crude (%)	0119	18.365	0.09000	18.610	0.31768	0.19568	195	-0.77	002.06	0
002	Protein, Crude (%)	0038	18.367	0.62100	18.610	0.31768	0.19568	195	-0.77	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein, Crude (%)	0309	18.375	0.25800	18.610	0.31768	0.19568	195	-0.74	002.06	0
002	Protein, Crude (%)	0897	18.380	0.02000	18.610	0.31768	0.19568	195	-0.72	002.02	0
002	Protein, Crude (%)	0148	18.380	0.30000	18.610	0.31768	0.19568	195	-0.72	002.06	0
002	Protein, Crude (%)	0964	18.387	0.34300	18.610	0.31768	0.19568	195	-0.70	002.06	0
002	Protein, Crude (%)	0848	18.390	0.07800	18.610	0.31768	0.19568	195	-0.69	002.06	0
002	Protein, Crude (%)	2195	18.405	0.09000	18.610	0.31768	0.19568	195	-0.65	002.01	0
002	Protein, Crude (%)	2022	18.405	0.13000	18.610	0.31768	0.19568	195	-0.65	002.05	0
002	Protein, Crude (%)	0968	18.410	0.02000	18.610	0.31768	0.19568	195	-0.63	002.01	0
002	Protein, Crude (%)	0066	18.410	0.10000	18.610	0.31768	0.19568	195	-0.63	002.02	0
002	Protein, Crude (%)	0720	18.415	0.03000	18.610	0.31768	0.19568	195	-0.61	002.06	0
002	Protein, Crude (%)	0878	18.420	0.04000	18.610	0.31768	0.19568	195	-0.60	002.05	0
002	Protein, Crude (%)	0881	18.420	0.04000	18.610	0.31768	0.19568	195	-0.60	002.05	0
002	Protein, Crude (%)	0049	18.420	0.48000	18.610	0.31768	0.19568	195	-0.60	002.06	0
002	Protein, Crude (%)	0032	18.422	0.11900	18.610	0.31768	0.19568	195	-0.59	002.06	0
002	Protein, Crude (%)	0047	18.425	0.05000	18.610	0.31768	0.19568	195	-0.58	002.06	0
002	Protein, Crude (%)	0139	18.425	0.03000	18.610	0.31768	0.19568	195	-0.58	002.06	0
002	Protein, Crude (%)	2196	18.450	0.00000	18.610	0.31768	0.19568	195	-0.50	002.01	0
002	Protein, Crude (%)	0849	18.454	0.06700	18.610	0.31768	0.19568	195	-0.49	002.05	0
002	Protein, Crude (%)	0354	18.455	0.01000	18.610	0.31768	0.19568	195	-0.49	002.05	0
002	Protein, Crude (%)	0563	18.456	0.17900	18.610	0.31768	0.19568	195	-0.49	002.08	0
002	Protein, Crude (%)	0857	18.460	0.18000	18.610	0.31768	0.19568	195	-0.47	002.06	0
002	Protein, Crude (%)	0870	18.463	0.23400	18.610	0.31768	0.19568	195	-0.46	002.01	0
002	Protein, Crude (%)	2150	18.467	0.78700	18.610	0.31768	0.19568	195	-0.45	002.06	0
002	Protein, Crude (%)	2168	18.470	0.12000	18.610	0.31768	0.19568	195	-0.44	002.06	0
002	Protein, Crude (%)	0870	18.475	0.19380	18.610	0.31768	0.19568	195	-0.43	002.06	0
002	Protein, Crude (%)	0034	18.495	0.06600	18.610	0.31768	0.19568	195	-0.36	002.06	0
002	Protein, Crude (%)	0505	18.495	0.31000	18.610	0.31768	0.19568	195	-0.36	002.06	0
002	Protein, Crude (%)	0692	18.500	1.4000	18.610	0.31768	0.19568	195	-0.35	002.06	0
002	Protein, Crude (%)	2193	18.505	0.17000	18.610	0.31768	0.19568	195	-0.33	002.01	0
002	Protein, Crude (%)	0042	18.505	0.19000	18.610	0.31768	0.19568	195	-0.33	002.02	0
002	Protein, Crude (%)	0939	18.510	0.30000	18.610	0.31768	0.19568	195	-0.32	002.05	0
002	Protein, Crude (%)	0504	18.525	0.15000	18.610	0.31768	0.19568	195	-0.27	002.04	0
002	Protein, Crude (%)	0074	18.525	0.17000	18.610	0.31768	0.19568	195	-0.27	002.06	0
002	Protein, Crude (%)	0861	18.525	0.09000	18.610	0.31768	0.19568	195	-0.27	002.06	0
002	Protein, Crude (%)	0722	18.537	0.11370	18.610	0.31768	0.19568	195	-0.23	002.05	0
002	Protein, Crude (%)	2126	18.540	0.06000	18.610	0.31768	0.19568	195	-0.22	002.06	0
002	Protein, Crude (%)	0504	18.542	0.19900	18.610	0.31768	0.19568	195	-0.22	002.06	0
002	Protein, Crude (%)	0019	18.545	0.13000	18.610	0.31768	0.19568	195	-0.21	002.06	0
002	Protein, Crude (%)	0868	18.550	0.02000	18.610	0.31768	0.19568	195	-0.19	002.06	0
002	Protein, Crude (%)	0689	18.550	0.10000	18.610	0.31768	0.19568	195	-0.19	002.05	0
002	Protein, Crude (%)	0009	18.550	0.10000	18.610	0.31768	0.19568	195	-0.19	002.06	0
002	Protein, Crude (%)	0017	18.550	0.10000	18.610	0.31768	0.19568	195	-0.19	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein, Crude (%)	0175	18.550	0.10000	18.610	0.31768	0.19568	195	-0.19	002.06	0
002	Protein, Crude (%)	0884	18.550	0.10000	18.610	0.31768	0.19568	195	-0.19	002.06	0
002	Protein, Crude (%)	0910	18.550	0.10000	18.610	0.31768	0.19568	195	-0.19	002.06	0
002	Protein, Crude (%)	0868	18.555	0.01000	18.610	0.31768	0.19568	195	-0.17	002.05	0
002	Protein, Crude (%)	0948	18.555	0.13000	18.610	0.31768	0.19568	195	-0.17	002.05	0
002	Protein, Crude (%)	0242	18.555	0.19000	18.610	0.31768	0.19568	195	-0.17	002.06	0
002	Protein, Crude (%)	0297	18.555	0.09000	18.610	0.31768	0.19568	195	-0.17	002.06	0
002	Protein, Crude (%)	0683	18.565	0.21000	18.610	0.31768	0.19568	195	-0.14	002.05	0
002	Protein, Crude (%)	0675	18.575	0.01000	18.610	0.31768	0.19568	195	-0.11	002.05	0
002	Protein, Crude (%)	2144	18.575	0.11000	18.610	0.31768	0.19568	195	-0.11	002.99	0
002	Protein, Crude (%)	2073	18.580	0.40000	18.610	0.31768	0.19568	195	-0.10	002.05	0
002	Protein, Crude (%)	0939	18.595	0.01000	18.610	0.31768	0.19568	195	-0.05	002.06	0
002	Protein, Crude (%)	0036	18.599	0.06390	18.610	0.31768	0.19568	195	-0.03	002.02	0
002	Protein, Crude (%)	0033	18.600	0.60000	18.610	0.31768	0.19568	195	-0.03	002.06	0
002	Protein, Crude (%)	0294	18.600	0.40000	18.610	0.31768	0.19568	195	-0.03	002.06	0
002	Protein, Crude (%)	0650	18.600	0.02000	18.610	0.31768	0.19568	195	-0.03	002.06	0
002	Protein, Crude (%)	0872	18.603	0.41900	18.610	0.31768	0.19568	195	-0.02	002.06	0
002	Protein, Crude (%)	0941	18.605	0.37000	18.610	0.31768	0.19568	195	-0.02	002.05	0
002	Protein, Crude (%)	0358	18.605	0.01000	18.610	0.31768	0.19568	195	-0.02	002.06	0
002	Protein, Crude (%)	2006	18.610	0.22000	18.610	0.31768	0.19568	195	0.00	002.05	0
002	Protein, Crude (%)	0003	18.615	0.23000	18.610	0.31768	0.19568	195	0.01	002.06	0
002	Protein, Crude (%)	0571	18.630	0.00000	18.610	0.31768	0.19568	195	0.06	002.06	0
002	Protein, Crude (%)	0948	18.630	0.20000	18.610	0.31768	0.19568	195	0.06	002.06	0
002	Protein, Crude (%)	0004	18.635	0.55000	18.610	0.31768	0.19568	195	0.08	002.06	0
002	Protein, Crude (%)	0144	18.635	0.09000	18.610	0.31768	0.19568	195	0.08	002.06	0
002	Protein, Crude (%)	0940	18.645	0.57000	18.610	0.31768	0.19568	195	0.11	002.00	0
002	Protein, Crude (%)	0089	18.650	0.00000	18.610	0.31768	0.19568	195	0.13	002.06	0
002	Protein, Crude (%)	0098	18.650	0.10000	18.610	0.31768	0.19568	195	0.13	002.06	0
002	Protein, Crude (%)	0590	18.650	0.10000	18.610	0.31768	0.19568	195	0.13	002.06	0
002	Protein, Crude (%)	0693	18.650	0.10000	18.610	0.31768	0.19568	195	0.13	002.06	0
002	Protein, Crude (%)	0298	18.660	0.02000	18.610	0.31768	0.19568	195	0.16	002.06	0
002	Protein, Crude (%)	0354	18.660	0.02000	18.610	0.31768	0.19568	195	0.16	002.06	0
002	Protein, Crude (%)	0682	18.670	0.00000	18.610	0.31768	0.19568	195	0.19	002.06	0
002	Protein, Crude (%)	0027	18.674	0.03700	18.610	0.31768	0.19568	195	0.20	002.06	0
002	Protein, Crude (%)	0407	18.687	0.03400	18.610	0.31768	0.19568	195	0.24	002.06	0
002	Protein, Crude (%)	0618	18.688	0.10600	18.610	0.31768	0.19568	195	0.24	002.06	0
002	Protein, Crude (%)	0205	18.690	0.02000	18.610	0.31768	0.19568	195	0.25	002.06	0
002	Protein, Crude (%)	0045	18.700	0.40000	18.610	0.31768	0.19568	195	0.28	002.06	0
002	Protein, Crude (%)	2161	18.706	0.03200	18.610	0.31768	0.19568	195	0.30	002.06	0
002	Protein, Crude (%)	0265	18.710	0.16000	18.610	0.31768	0.19568	195	0.31	002.06	0
002	Protein, Crude (%)	0037	18.720	0.30000	18.610	0.31768	0.19568	195	0.35	002.06	0
002	Protein, Crude (%)	0574	18.740	0.14000	18.610	0.31768	0.19568	195	0.41	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein, Crude (%)	0563	18.748	0.25500	18.610	0.31768	0.19568	195	0.43	002.06	0
002	Protein, Crude (%)	0108	18.750	0.22000	18.610	0.31768	0.19568	195	0.44	002.06	0
002	Protein, Crude (%)	0208	18.750	0.10000	18.610	0.31768	0.19568	195	0.44	002.06	0
002	Protein, Crude (%)	0278	18.750	0.10000	18.610	0.31768	0.19568	195	0.44	002.06	0
002	Protein, Crude (%)	0357	18.750	0.25000	18.610	0.31768	0.19568	195	0.44	002.06	0
002	Protein, Crude (%)	0589	18.755	0.01000	18.610	0.31768	0.19568	195	0.46	002.06	0
002	Protein, Crude (%)	0300	18.760	0.36000	18.610	0.31768	0.19568	195	0.47	002.06	0
002	Protein, Crude (%)	2053	18.760	0.20000	18.610	0.31768	0.19568	195	0.47	002.06	0
002	Protein, Crude (%)	0626	18.761	0.00200	18.610	0.31768	0.19568	195	0.47	002.06	0
002	Protein, Crude (%)	0035	18.765	0.05000	18.610	0.31768	0.19568	195	0.49	002.06	0
002	Protein, Crude (%)	2089	18.770	0.02000	18.610	0.31768	0.19568	195	0.50	002.06	0
002	Protein, Crude (%)	0541	18.785	0.31000	18.610	0.31768	0.19568	195	0.55	002.06	0
002	Protein, Crude (%)	0683	18.790	0.24000	18.610	0.31768	0.19568	195	0.57	002.06	0
002	Protein, Crude (%)	2076	18.796	0.13200	18.610	0.31768	0.19568	195	0.58	002.06	0
002	Protein, Crude (%)	0652	18.800	0.20000	18.610	0.31768	0.19568	195	0.60	002.01	0
002	Protein, Crude (%)	2200	18.800	0.20000	18.610	0.31768	0.19568	195	0.60	002.05	0
002	Protein, Crude (%)	0010	18.800	0.00000	18.610	0.31768	0.19568	195	0.60	002.06	0
002	Protein, Crude (%)	0675	18.815	0.03000	18.610	0.31768	0.19568	195	0.64	002.06	0
002	Protein, Crude (%)	2190	18.820	0.06000	18.610	0.31768	0.19568	195	0.66	002.06	0
002	Protein, Crude (%)	0190	18.825	0.03000	18.610	0.31768	0.19568	195	0.68	002.06	0
002	Protein, Crude (%)	0028	18.825	0.41000	18.610	0.31768	0.19568	195	0.68	002.06	0
002	Protein, Crude (%)	0941	18.825	0.23000	18.610	0.31768	0.19568	195	0.68	002.06	0
002	Protein, Crude (%)	0726	18.835	0.01000	18.610	0.31768	0.19568	195	0.71	002.06	0
002	Protein, Crude (%)	0263	18.844	0.00870	18.610	0.31768	0.19568	195	0.74	002.06	0
002	Protein, Crude (%)	2129	18.846	0.37200	18.610	0.31768	0.19568	195	0.74	002.06	0
002	Protein, Crude (%)	0013	18.860	0.16000	18.610	0.31768	0.19568	195	0.79	002.06	0
002	Protein, Crude (%)	0229	18.860	0.06000	18.610	0.31768	0.19568	195	0.79	002.06	0
002	Protein, Crude (%)	0843	18.865	0.63000	18.610	0.31768	0.19568	195	0.80	002.06	0
002	Protein, Crude (%)	0644	18.868	0.15700	18.610	0.31768	0.19568	195	0.81	002.06	0
002	Protein, Crude (%)	0042	18.870	0.14000	18.610	0.31768	0.19568	195	0.82	002.06	0
002	Protein, Crude (%)	0152	18.875	0.09000	18.610	0.31768	0.19568	195	0.83	002.01	0
002	Protein, Crude (%)	0968	18.875	0.22600	18.610	0.31768	0.19568	195	0.83	002.06	0
002	Protein, Crude (%)	0918	18.885	0.75000	18.610	0.31768	0.19568	195	0.86	002.06	0
002	Protein, Crude (%)	2006	18.895	0.19000	18.610	0.31768	0.19568	195	0.90	002.06	0
002	Protein, Crude (%)	2199	18.900	0.20000	18.610	0.31768	0.19568	195	0.91	002.05	0
002	Protein, Crude (%)	0598	18.900	0.02000	18.610	0.31768	0.19568	195	0.91	002.06	0
002	Protein, Crude (%)	0687	18.900	0.20000	18.610	0.31768	0.19568	195	0.91	002.06	0
002	Protein, Crude (%)	0596	18.930	0.28000	18.610	0.31768	0.19568	195	1.01	002.06	0
002	Protein, Crude (%)	0353	18.950	0.26000	18.610	0.31768	0.19568	195	1.07	002.06	0
002	Protein, Crude (%)	0425	18.950	0.14000	18.610	0.31768	0.19568	195	1.07	002.06	0
002	Protein, Crude (%)	0142	18.955	0.07000	18.610	0.31768	0.19568	195	1.09	002.06	0
002	Protein, Crude (%)	0630	18.955	0.01000	18.610	0.31768	0.19568	195	1.09	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein, Crude (%)	0164	18.970	0.24000	18.610	0.31768	0.19568	195	1.13	002.06	0
002	Protein, Crude (%)	0511	18.980	0.14000	18.610	0.31768	0.19568	195	1.16	002.06	0
002	Protein, Crude (%)	0186	19.000	0.00000	18.610	0.31768	0.19568	195	1.23	002.06	0
002	Protein, Crude (%)	0609	19.030	0.06000	18.610	0.31768	0.19568	195	1.32	002.06	0
002	Protein, Crude (%)	0553	19.055	1.0700	18.610	0.31768	0.19568	195	1.40	002.06	0
002	Protein, Crude (%)	0062	19.070	0.05300	18.610	0.31768	0.19568	195	1.45	002.06	0
002	Protein, Crude (%)	2145	19.080	0.02000	18.610	0.31768	0.19568	195	1.48	002.05	0
002	Protein, Crude (%)	0610	19.100	0.10000	18.610	0.31768	0.19568	195	1.54	002.01	0
002	Protein, Crude (%)	0345	19.100	0.40000	18.610	0.31768	0.19568	195	1.54	002.06	0
002	Protein, Crude (%)	2059	19.111	0.07180	18.610	0.31768	0.19568	195	1.58	002.06	0
002	Protein, Crude (%)	0610	19.125	0.25000	18.610	0.31768	0.19568	195	1.62	002.06	0
002	Protein, Crude (%)	0001	19.150	0.18000	18.610	0.31768	0.19568	195	1.70	002.06	0
002	Protein, Crude (%)	0520	19.150	0.40000	18.610	0.31768	0.19568	195	1.70	002.06	0
002	Protein, Crude (%)	0508	19.195	0.26400	18.610	0.31768	0.19568	195	1.84	002.06	0
002	Protein, Crude (%)	0859	19.248	0.37400	18.610	0.31768	0.19568	195	2.01	002.06	0
002	Protein, Crude (%)	0036	19.260	0.12000	18.610	0.31768	0.19568	195	2.05	002.06	0
002	Protein, Crude (%)	2057	19.435	0.47000	18.610	0.31768	0.19568	195	2.60	002.06	0
002	Protein, Crude (%)	2190	20.065	0.05000	18.610	0.31768	0.19568	195	4.58	002.11	0
002	Protein, Crude (%)	0953	20.090	0.66000	18.610	0.31768	0.19568	195	4.66	002.06	0
002	Protein, Crude (%)	0852	20.735	0.11000	18.610	0.31768	0.19568	195	6.69	002.11	0
002	Protein, Crude (%)	2128	21.209	0.14300	18.610	0.31768	0.19568	195	8.18	002.11	0
002	Protein, Crude (%)	0297	21.210	0.10000	18.610	0.31768	0.19568	195	8.18	002.11	0
002	Protein, Crude (%)	2120	21.540	0.38000	18.610	0.31768	0.19568	195	9.22	002.11	0
002	Protein, Crude (%)	2121	21.750	0.20000	18.610	0.31768	0.19568	195	9.88	002.11	0
002	Protein, Crude (%)	0405	22.285	0.29000	18.610	0.31768	0.19568	195	11.57	002.04	0
002	Protein, Crude (%)	0720	24.010	0.04000	18.610	0.31768	0.19568	195	17.00	002.11	0
002	Protein, Crude (%)	0510	3.0050	0.03000	18.610	0.31768	0.19568	195	-49.12	002.06	2
002	Protein, Crude (%)	0889	33.735	29.910	18.610	0.31768	0.19568	195	47.61	002.01	2
003	Fat, Crude (%)	0848	2.0000	0.02000	3.3223	0.49142	0.09617	150	-2.69	003.14	0
003	Fat, Crude (%)	0897	2.1850	0.01000	3.3223	0.49142	0.09617	150	-2.31	003.10	0
003	Fat, Crude (%)	2120	2.1900	0.12000	3.3223	0.49142	0.09617	150	-2.30	003.11	0
003	Fat, Crude (%)	0968	2.2675	0.13700	3.3223	0.49142	0.09617	150	-2.15	003.14	0
003	Fat, Crude (%)	0265	2.3300	0.08000	3.3223	0.49142	0.09617	150	-2.02	003.14	0
003	Fat, Crude (%)	0912	2.3550	0.01000	3.3223	0.49142	0.09617	150	-1.97	003.14	0
003	Fat, Crude (%)	0186	2.3800	0.06000	3.3223	0.49142	0.09617	150	-1.92	003.14	0
003	Fat, Crude (%)	0019	2.4150	0.17000	3.3223	0.49142	0.09617	150	-1.85	003.14	0
003	Fat, Crude (%)	2129	2.4400	0.10000	3.3223	0.49142	0.09617	150	-1.80	003.14	0
003	Fat, Crude (%)	0142	2.4950	0.25000	3.3223	0.49142	0.09617	150	-1.68	003.00	0
003	Fat, Crude (%)	2076	2.5292	0.27740	3.3223	0.49142	0.09617	150	-1.61	003.14	0
003	Fat, Crude (%)	0175	2.5500	0.06000	3.3223	0.49142	0.09617	150	-1.57	003.14	0
003	Fat, Crude (%)	0581	2.5500	0.16000	3.3223	0.49142	0.09617	150	-1.57	003.14	0
003	Fat, Crude (%)	0952	2.5800	0.02000	3.3223	0.49142	0.09617	150	-1.51	003.10	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat, Crude (%)	0955	2.6150	0.03000	3.3223	0.49142	0.09617	150	-1.44	003.14	0
003	Fat, Crude (%)	2155	2.6190	0.09000	3.3223	0.49142	0.09617	150	-1.43	003.10	0
003	Fat, Crude (%)	2121	2.6300	0.02000	3.3223	0.49142	0.09617	150	-1.41	003.11	0
003	Fat, Crude (%)	0358	2.6400	0.02000	3.3223	0.49142	0.09617	150	-1.39	003.14	0
003	Fat, Crude (%)	0536	2.6450	0.07000	3.3223	0.49142	0.09617	150	-1.38	003.99	0
003	Fat, Crude (%)	0843	2.6600	0.26000	3.3223	0.49142	0.09617	150	-1.35	003.14	0
003	Fat, Crude (%)	0144	2.6750	0.09000	3.3223	0.49142	0.09617	150	-1.32	003.14	0
003	Fat, Crude (%)	0089	2.7100	0.00000	3.3223	0.49142	0.09617	150	-1.25	003.14	0
003	Fat, Crude (%)	0037	2.7250	0.03000	3.3223	0.49142	0.09617	150	-1.22	003.14	0
003	Fat, Crude (%)	0885	2.7496	0.01080	3.3223	0.49142	0.09617	150	-1.17	003.00	0
003	Fat, Crude (%)	0357	2.7500	0.30000	3.3223	0.49142	0.09617	150	-1.16	003.14	0
003	Fat, Crude (%)	0574	2.7500	0.02000	3.3223	0.49142	0.09617	150	-1.16	003.14	0
003	Fat, Crude (%)	0598	2.7850	0.11000	3.3223	0.49142	0.09617	150	-1.09	003.14	0
003	Fat, Crude (%)	0692	2.8000	0.40000	3.3223	0.49142	0.09617	150	-1.06	003.14	0
003	Fat, Crude (%)	0861	2.8150	0.05000	3.3223	0.49142	0.09617	150	-1.03	003.10	0
003	Fat, Crude (%)	0878	2.8300	0.10000	3.3223	0.49142	0.09617	150	-1.00	003.14	0
003	Fat, Crude (%)	0407	2.8328	0.20810	3.3223	0.49142	0.09617	150	-1.00	003.14	0
003	Fat, Crude (%)	0520	2.8500	0.02000	3.3223	0.49142	0.09617	150	-0.96	003.14	0
003	Fat, Crude (%)	0956	2.8500	0.10000	3.3223	0.49142	0.09617	150	-0.96	003.14	0
003	Fat, Crude (%)	0049	2.8900	0.10000	3.3223	0.49142	0.09617	150	-0.88	003.14	0
003	Fat, Crude (%)	0683	2.9050	0.03000	3.3223	0.49142	0.09617	150	-0.85	003.14	0
003	Fat, Crude (%)	0939	2.9600	0.16000	3.3223	0.49142	0.09617	150	-0.74	003.00	0
003	Fat, Crude (%)	0868	2.9950	0.01000	3.3223	0.49142	0.09617	150	-0.67	003.10	0
003	Fat, Crude (%)	0263	3.0000	0.00000	3.3223	0.49142	0.09617	150	-0.66	003.09	0
003	Fat, Crude (%)	0013	3.0000	0.20000	3.3223	0.49142	0.09617	150	-0.66	003.14	0
003	Fat, Crude (%)	0278	3.0000	0.40000	3.3223	0.49142	0.09617	150	-0.66	003.14	0
003	Fat, Crude (%)	0001	3.0015	0.07700	3.3223	0.49142	0.09617	150	-0.65	003.14	0
003	Fat, Crude (%)	0027	3.0260	0.27000	3.3223	0.49142	0.09617	150	-0.60	003.09	0
003	Fat, Crude (%)	0098	3.0400	0.04000	3.3223	0.49142	0.09617	150	-0.57	003.10	0
003	Fat, Crude (%)	0720	3.0450	0.03000	3.3223	0.49142	0.09617	150	-0.56	003.11	0
003	Fat, Crude (%)	0229	3.0675	0.15300	3.3223	0.49142	0.09617	150	-0.52	003.14	0
003	Fat, Crude (%)	0651	3.0945	0.19700	3.3223	0.49142	0.09617	150	-0.46	003.10	0
003	Fat, Crude (%)	0034	3.1200	0.40000	3.3223	0.49142	0.09617	150	-0.41	003.14	0
003	Fat, Crude (%)	0948	3.1300	0.02000	3.3223	0.49142	0.09617	150	-0.39	003.14	0
003	Fat, Crude (%)	0563	3.1625	0.04450	3.3223	0.49142	0.09617	150	-0.33	003.06	0
003	Fat, Crude (%)	0013	3.1750	0.29000	3.3223	0.49142	0.09617	150	-0.30	003.09	0
003	Fat, Crude (%)	0042	3.2050	0.19000	3.3223	0.49142	0.09617	150	-0.24	003.10	0
003	Fat, Crude (%)	2089	3.2100	0.02000	3.3223	0.49142	0.09617	150	-0.23	003.13	0
003	Fat, Crude (%)	2073	3.2150	0.07000	3.3223	0.49142	0.09617	150	-0.22	003.10	0
003	Fat, Crude (%)	2022	3.2300	0.02000	3.3223	0.49142	0.09617	150	-0.19	003.10	0
003	Fat, Crude (%)	0066	3.2400	0.14000	3.3223	0.49142	0.09617	150	-0.17	003.14	0
003	Fat, Crude (%)	2196	3.2600	0.00000	3.3223	0.49142	0.09617	150	-0.13	003.10	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat, Crude (%)	0553	3.2750	0.23000	3.3223	0.49142	0.09617	150	-0.10	003.10	0
003	Fat, Crude (%)	0618	3.2885	0.13700	3.3223	0.49142	0.09617	150	-0.07	003.10	0
003	Fat, Crude (%)	2190	3.2900	0.14000	3.3223	0.49142	0.09617	150	-0.07	003.00	0
003	Fat, Crude (%)	2188	3.3000	0.02000	3.3223	0.49142	0.09617	150	-0.05	003.10	0
003	Fat, Crude (%)	2193	3.3050	0.01000	3.3223	0.49142	0.09617	150	-0.04	003.10	0
003	Fat, Crude (%)	0298	3.3100	0.02000	3.3223	0.49142	0.09617	150	-0.03	003.10	0
003	Fat, Crude (%)	0009	3.3100	0.04000	3.3223	0.49142	0.09617	150	-0.03	003.14	0
003	Fat, Crude (%)	0242	3.3150	0.03000	3.3223	0.49142	0.09617	150	-0.01	003.10	0
003	Fat, Crude (%)	0171	3.3300	0.06000	3.3223	0.49142	0.09617	150	0.02	003.12	0
003	Fat, Crude (%)	0074	3.3400	0.10000	3.3223	0.49142	0.09617	150	0.04	003.14	0
003	Fat, Crude (%)	0062	3.3600	0.03000	3.3223	0.49142	0.09617	150	0.08	003.10	0
003	Fat, Crude (%)	0047	3.3650	0.01000	3.3223	0.49142	0.09617	150	0.09	003.99	0
003	Fat, Crude (%)	0940	3.3700	0.44000	3.3223	0.49142	0.09617	150	0.10	003.12	0
003	Fat, Crude (%)	0036	3.3745	0.05880	3.3223	0.49142	0.09617	150	0.11	003.10	0
003	Fat, Crude (%)	0895	3.3750	0.01000	3.3223	0.49142	0.09617	150	0.11	003.12	0
003	Fat, Crude (%)	0609	3.3800	0.12000	3.3223	0.49142	0.09617	150	0.12	003.10	0
003	Fat, Crude (%)	0033	3.3800	0.02000	3.3223	0.49142	0.09617	150	0.12	003.13	0
003	Fat, Crude (%)	0004	3.3850	0.15000	3.3223	0.49142	0.09617	150	0.13	003.14	0
003	Fat, Crude (%)	0504	3.3900	0.06000	3.3223	0.49142	0.09617	150	0.14	003.01	0
003	Fat, Crude (%)	0208	3.3900	0.06000	3.3223	0.49142	0.09617	150	0.14	003.13	0
003	Fat, Crude (%)	0045	3.3950	0.21000	3.3223	0.49142	0.09617	150	0.15	003.10	0
003	Fat, Crude (%)	0720	3.3950	0.29000	3.3223	0.49142	0.09617	150	0.15	003.10	0
003	Fat, Crude (%)	0722	3.3995	0.05660	3.3223	0.49142	0.09617	150	0.16	003.09	0
003	Fat, Crude (%)	0675	3.4100	0.04000	3.3223	0.49142	0.09617	150	0.18	003.09	0
003	Fat, Crude (%)	0229	3.4240	0.03600	3.3223	0.49142	0.09617	150	0.21	003.06	0
003	Fat, Crude (%)	0693	3.4500	0.10000	3.3223	0.49142	0.09617	150	0.26	003.10	0
003	Fat, Crude (%)	2195	3.4500	0.02000	3.3223	0.49142	0.09617	150	0.26	003.10	0
003	Fat, Crude (%)	0202	3.4550	0.13000	3.3223	0.49142	0.09617	150	0.27	003.14	0
003	Fat, Crude (%)	2009	3.4724	0.00640	3.3223	0.49142	0.09617	150	0.31	003.10	0
003	Fat, Crude (%)	0169	3.4750	0.03000	3.3223	0.49142	0.09617	150	0.31	003.06	0
003	Fat, Crude (%)	0948	3.4750	0.01000	3.3223	0.49142	0.09617	150	0.31	003.09	0
003	Fat, Crude (%)	0297	3.4800	0.10000	3.3223	0.49142	0.09617	150	0.32	003.06	0
003	Fat, Crude (%)	0682	3.4800	0.00000	3.3223	0.49142	0.09617	150	0.32	003.14	0
003	Fat, Crude (%)	2192	3.4800	0.02000	3.3223	0.49142	0.09617	150	0.32	003.00	0
003	Fat, Crude (%)	0629	3.4900	0.06000	3.3223	0.49142	0.09617	150	0.34	003.10	0
003	Fat, Crude (%)	0015	3.4900	0.22000	3.3223	0.49142	0.09617	150	0.34	003.14	0
003	Fat, Crude (%)	2146	3.4950	0.17000	3.3223	0.49142	0.09617	150	0.35	003.14	0
003	Fat, Crude (%)	0083	3.5250	0.23000	3.3223	0.49142	0.09617	150	0.41	003.06	0
003	Fat, Crude (%)	0035	3.5400	0.04000	3.3223	0.49142	0.09617	150	0.44	003.00	0
003	Fat, Crude (%)	2128	3.5456	0.03710	3.3223	0.49142	0.09617	150	0.45	003.11	0
003	Fat, Crude (%)	0644	3.5475	0.05100	3.3223	0.49142	0.09617	150	0.46	003.06	0
003	Fat, Crude (%)	0870	3.5476	0.00350	3.3223	0.49142	0.09617	150	0.46	003.10	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat, Crude (%)	0345	3.5750	0.05000	3.3223	0.49142	0.09617	150	0.51	003.00	0
003	Fat, Crude (%)	0590	3.5750	0.33000	3.3223	0.49142	0.09617	150	0.51	003.09	0
003	Fat, Crude (%)	0187	3.5800	0.06000	3.3223	0.49142	0.09617	150	0.52	003.13	0
003	Fat, Crude (%)	0689	3.6000	0.20000	3.3223	0.49142	0.09617	150	0.57	003.10	0
003	Fat, Crude (%)	2200	3.6000	0.00000	3.3223	0.49142	0.09617	150	0.57	003.12	0
003	Fat, Crude (%)	0205	3.6100	0.16000	3.3223	0.49142	0.09617	150	0.59	003.13	0
003	Fat, Crude (%)	0108	3.6150	0.29000	3.3223	0.49142	0.09617	150	0.60	003.14	0
003	Fat, Crude (%)	0309	3.6182	0.17310	3.3223	0.49142	0.09617	150	0.60	003.00	0
003	Fat, Crude (%)	2181	3.6200	0.06000	3.3223	0.49142	0.09617	150	0.61	003.06	0
003	Fat, Crude (%)	0100	3.6200	0.18000	3.3223	0.49142	0.09617	150	0.61	003.10	0
003	Fat, Crude (%)	0098	3.6300	0.02000	3.3223	0.49142	0.09617	150	0.63	003.09	0
003	Fat, Crude (%)	0849	3.6350	0.05000	3.3223	0.49142	0.09617	150	0.64	003.06	0
003	Fat, Crude (%)	0676	3.6500	0.08000	3.3223	0.49142	0.09617	150	0.67	003.10	0
003	Fat, Crude (%)	0884	3.6500	0.10000	3.3223	0.49142	0.09617	150	0.67	003.06	0
003	Fat, Crude (%)	0413	3.6500	0.10000	3.3223	0.49142	0.09617	150	0.67	003.14	0
003	Fat, Crude (%)	0152	3.6700	0.14000	3.3223	0.49142	0.09617	150	0.71	003.00	0
003	Fat, Crude (%)	0148	3.6700	0.00000	3.3223	0.49142	0.09617	150	0.71	003.06	0
003	Fat, Crude (%)	0689	3.7000	0.00000	3.3223	0.49142	0.09617	150	0.77	003.06	0
003	Fat, Crude (%)	0004	3.7000	0.00000	3.3223	0.49142	0.09617	150	0.77	003.09	0
003	Fat, Crude (%)	0226	3.7000	0.00000	3.3223	0.49142	0.09617	150	0.77	003.09	0
003	Fat, Crude (%)	0353	3.7050	0.19000	3.3223	0.49142	0.09617	150	0.78	003.09	0
003	Fat, Crude (%)	2144	3.7100	0.02000	3.3223	0.49142	0.09617	150	0.79	003.99	0
003	Fat, Crude (%)	0541	3.7150	0.03000	3.3223	0.49142	0.09617	150	0.80	003.10	0
003	Fat, Crude (%)	0889	3.7300	0.02000	3.3223	0.49142	0.09617	150	0.83	003.10	0
003	Fat, Crude (%)	0726	3.7400	0.02000	3.3223	0.49142	0.09617	150	0.85	003.14	0
003	Fat, Crude (%)	0164	3.7450	0.09000	3.3223	0.49142	0.09617	150	0.86	003.06	0
003	Fat, Crude (%)	0630	3.7550	0.01000	3.3223	0.49142	0.09617	150	0.88	003.09	0
003	Fat, Crude (%)	0508	3.7643	0.01360	3.3223	0.49142	0.09617	150	0.90	003.09	0
003	Fat, Crude (%)	0175	3.7650	0.03000	3.3223	0.49142	0.09617	150	0.90	003.00	0
003	Fat, Crude (%)	0941	3.7650	0.21000	3.3223	0.49142	0.09617	150	0.90	003.99	0
003	Fat, Crude (%)	2006	3.7700	0.04000	3.3223	0.49142	0.09617	150	0.91	003.13	0
003	Fat, Crude (%)	0653	3.7750	0.05000	3.3223	0.49142	0.09617	150	0.92	003.00	0
003	Fat, Crude (%)	0032	3.7750	0.27000	3.3223	0.49142	0.09617	150	0.92	003.00	0
003	Fat, Crude (%)	0425	3.7800	0.04000	3.3223	0.49142	0.09617	150	0.93	003.06	0
003	Fat, Crude (%)	2109	3.7800	0.02000	3.3223	0.49142	0.09617	150	0.93	003.00	0
003	Fat, Crude (%)	2053	3.7900	0.00000	3.3223	0.49142	0.09617	150	0.95	003.00	0
003	Fat, Crude (%)	0687	3.8000	0.20000	3.3223	0.49142	0.09617	150	0.97	003.06	0
003	Fat, Crude (%)	0910	3.8000	0.10000	3.3223	0.49142	0.09617	150	0.97	003.06	0
003	Fat, Crude (%)	0190	3.8400	0.24000	3.3223	0.49142	0.09617	150	1.05	003.14	0
003	Fat, Crude (%)	0510	3.8500	0.10000	3.3223	0.49142	0.09617	150	1.07	003.09	0
003	Fat, Crude (%)	0881	3.8560	0.02800	3.3223	0.49142	0.09617	150	1.09	003.10	0
003	Fat, Crude (%)	0852	3.8950	0.03000	3.3223	0.49142	0.09617	150	1.17	003.11	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat, Crude (%)	0354	3.9250	0.01000	3.3223	0.49142	0.09617	150	1.23	003.09	0
003	Fat, Crude (%)	0685	3.9300	0.00000	3.3223	0.49142	0.09617	150	1.24	003.09	0
003	Fat, Crude (%)	0682	3.9800	0.00000	3.3223	0.49142	0.09617	150	1.34	003.06	0
003	Fat, Crude (%)	0918	4.0450	0.03000	3.3223	0.49142	0.09617	150	1.47	003.06	0
003	Fat, Crude (%)	0017	4.0500	0.10000	3.3223	0.49142	0.09617	150	1.48	003.00	0
003	Fat, Crude (%)	2126	4.0700	0.02000	3.3223	0.49142	0.09617	150	1.52	003.00	0
003	Fat, Crude (%)	0511	4.1750	0.15000	3.3223	0.49142	0.09617	150	1.74	003.06	0
003	Fat, Crude (%)	0596	4.2550	0.13000	3.3223	0.49142	0.09617	150	1.90	003.00	0
003	Fat, Crude (%)	2161	4.3750	0.35000	3.3223	0.49142	0.09617	150	2.14	003.99	0
003	Fat, Crude (%)	0297	4.7800	0.02000	3.3223	0.49142	0.09617	150	2.97	003.11	0
003	Fat, Crude (%)	0003	2.0550	1.3900	3.3223	0.49142	0.09617	150	-2.58	003.14	1
003	Fat, Crude (%)	0294	2.9600	1.0200	3.3223	0.49142	0.09617	150	-0.74	003.06	1
003	Fat, Crude (%)	0889	6.7950	0.19000	3.3223	0.49142	0.09617	150	7.07	003.11	2
004	Fiber, Crude (%)	0889	2.4650	0.05000	3.8103	0.39167	0.16501	125	-3.43	004.11	0
004	Fiber, Crude (%)	2126	2.8200	0.30000	3.8103	0.39167	0.16501	125	-2.53	004.99	0
004	Fiber, Crude (%)	0169	3.0150	0.17000	3.8103	0.39167	0.16501	125	-2.03	004.00	0
004	Fiber, Crude (%)	2089	3.0600	0.24000	3.8103	0.39167	0.16501	125	-1.92	004.03	0
004	Fiber, Crude (%)	0921	3.1150	0.21000	3.8103	0.39167	0.16501	125	-1.78	004.03	0
004	Fiber, Crude (%)	0407	3.1509	0.02110	3.8103	0.39167	0.16501	125	-1.68	004.07	0
004	Fiber, Crude (%)	0028	3.2000	0.20000	3.8103	0.39167	0.16501	125	-1.56	004.07	0
004	Fiber, Crude (%)	0683	3.2300	0.04000	3.8103	0.39167	0.16501	125	-1.48	004.07	0
004	Fiber, Crude (%)	0074	3.2450	0.21000	3.8103	0.39167	0.16501	125	-1.44	004.07	0
004	Fiber, Crude (%)	2161	3.2950	0.43000	3.8103	0.39167	0.16501	125	-1.32	004.03	0
004	Fiber, Crude (%)	0536	3.3100	0.08000	3.8103	0.39167	0.16501	125	-1.28	004.07	0
004	Fiber, Crude (%)	2066	3.3350	0.51000	3.8103	0.39167	0.16501	125	-1.21	004.00	0
004	Fiber, Crude (%)	0910	3.3500	0.50000	3.8103	0.39167	0.16501	125	-1.18	004.07	0
004	Fiber, Crude (%)	2188	3.3550	0.01000	3.8103	0.39167	0.16501	125	-1.16	004.07	0
004	Fiber, Crude (%)	0098	3.3850	0.13000	3.8103	0.39167	0.16501	125	-1.09	004.07	0
004	Fiber, Crude (%)	0708	3.3850	0.13000	3.8103	0.39167	0.16501	125	-1.09	004.07	0
004	Fiber, Crude (%)	0309	3.4020	0.55160	3.8103	0.39167	0.16501	125	-1.04	004.00	0
004	Fiber, Crude (%)	0019	3.4150	0.17000	3.8103	0.39167	0.16501	125	-1.01	004.07	0
004	Fiber, Crude (%)	2146	3.4150	0.57000	3.8103	0.39167	0.16501	125	-1.01	004.07	0
004	Fiber, Crude (%)	0505	3.4300	0.12000	3.8103	0.39167	0.16501	125	-0.97	004.07	0
004	Fiber, Crude (%)	0013	3.4500	0.16000	3.8103	0.39167	0.16501	125	-0.92	004.07	0
004	Fiber, Crude (%)	0171	3.4750	0.07000	3.8103	0.39167	0.16501	125	-0.86	004.00	0
004	Fiber, Crude (%)	0042	3.4950	0.03000	3.8103	0.39167	0.16501	125	-0.80	004.07	0
004	Fiber, Crude (%)	2146	3.5000	0.02000	3.8103	0.39167	0.16501	125	-0.79	004.99	0
004	Fiber, Crude (%)	0885	3.5116	0.01710	3.8103	0.39167	0.16501	125	-0.76	004.06	0
004	Fiber, Crude (%)	2145	3.5150	0.05000	3.8103	0.39167	0.16501	125	-0.75	004.99	0
004	Fiber, Crude (%)	0720	3.5300	0.00000	3.8103	0.39167	0.16501	125	-0.72	004.06	0
004	Fiber, Crude (%)	0298	3.5350	0.07000	3.8103	0.39167	0.16501	125	-0.70	004.00	0
004	Fiber, Crude (%)	0425	3.5500	0.10000	3.8103	0.39167	0.16501	125	-0.66	004.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber, Crude (%)	0726	3.5500	0.10000	3.8103	0.39167	0.16501	125	-0.66	004.07	0
004	Fiber, Crude (%)	0098	3.5550	0.25000	3.8103	0.39167	0.16501	125	-0.65	004.06	0
004	Fiber, Crude (%)	0100	3.5700	0.60000	3.8103	0.39167	0.16501	125	-0.61	004.07	0
004	Fiber, Crude (%)	2076	3.5714	0.15780	3.8103	0.39167	0.16501	125	-0.61	004.00	0
004	Fiber, Crude (%)	2193	3.5750	0.45000	3.8103	0.39167	0.16501	125	-0.60	004.07	0
004	Fiber, Crude (%)	0590	3.5800	0.06000	3.8103	0.39167	0.16501	125	-0.59	004.06	0
004	Fiber, Crude (%)	0345	3.6000	0.10000	3.8103	0.39167	0.16501	125	-0.54	004.00	0
004	Fiber, Crude (%)	2200	3.6000	0.20000	3.8103	0.39167	0.16501	125	-0.54	004.06	0
004	Fiber, Crude (%)	0682	3.6000	0.00000	3.8103	0.39167	0.16501	125	-0.54	004.07	0
004	Fiber, Crude (%)	2006	3.6050	0.09000	3.8103	0.39167	0.16501	125	-0.52	004.06	0
004	Fiber, Crude (%)	0939	3.6100	0.38000	3.8103	0.39167	0.16501	125	-0.51	004.07	0
004	Fiber, Crude (%)	0353	3.6200	0.12000	3.8103	0.39167	0.16501	125	-0.49	004.03	0
004	Fiber, Crude (%)	2109	3.6250	0.17000	3.8103	0.39167	0.16501	125	-0.47	004.07	0
004	Fiber, Crude (%)	0918	3.6450	0.19000	3.8103	0.39167	0.16501	125	-0.42	004.06	0
004	Fiber, Crude (%)	0164	3.6500	0.10000	3.8103	0.39167	0.16501	125	-0.41	004.00	0
004	Fiber, Crude (%)	2073	3.6500	0.14000	3.8103	0.39167	0.16501	125	-0.41	004.00	0
004	Fiber, Crude (%)	0952	3.6550	0.01000	3.8103	0.39167	0.16501	125	-0.40	004.06	0
004	Fiber, Crude (%)	2022	3.6550	0.13000	3.8103	0.39167	0.16501	125	-0.40	004.06	0
004	Fiber, Crude (%)	0096	3.6600	0.04000	3.8103	0.39167	0.16501	125	-0.38	004.07	0
004	Fiber, Crude (%)	0354	3.6850	0.03000	3.8103	0.39167	0.16501	125	-0.32	004.07	0
004	Fiber, Crude (%)	0861	3.6850	0.21000	3.8103	0.39167	0.16501	125	-0.32	004.07	0
004	Fiber, Crude (%)	0940	3.6900	0.16000	3.8103	0.39167	0.16501	125	-0.31	004.99	0
004	Fiber, Crude (%)	0884	3.7000	0.00000	3.8103	0.39167	0.16501	125	-0.28	004.00	0
004	Fiber, Crude (%)	0653	3.7050	0.07000	3.8103	0.39167	0.16501	125	-0.27	004.07	0
004	Fiber, Crude (%)	0175	3.7150	0.33000	3.8103	0.39167	0.16501	125	-0.24	004.00	0
004	Fiber, Crude (%)	0205	3.7200	0.20000	3.8103	0.39167	0.16501	125	-0.23	004.06	0
004	Fiber, Crude (%)	0300	3.7200	0.18000	3.8103	0.39167	0.16501	125	-0.23	004.07	0
004	Fiber, Crude (%)	0013	3.7250	0.11000	3.8103	0.39167	0.16501	125	-0.22	004.00	0
004	Fiber, Crude (%)	0190	3.7500	0.30000	3.8103	0.39167	0.16501	125	-0.15	004.07	0
004	Fiber, Crude (%)	0510	3.7500	0.30000	3.8103	0.39167	0.16501	125	-0.15	004.07	0
004	Fiber, Crude (%)	0354	3.7700	0.02000	3.8103	0.39167	0.16501	125	-0.10	004.00	0
004	Fiber, Crude (%)	0033	3.7700	0.04000	3.8103	0.39167	0.16501	125	-0.10	004.07	0
004	Fiber, Crude (%)	0596	3.7800	0.30000	3.8103	0.39167	0.16501	125	-0.08	004.00	0
004	Fiber, Crude (%)	2144	3.7800	0.02000	3.8103	0.39167	0.16501	125	-0.08	004.99	0
004	Fiber, Crude (%)	0511	3.8000	0.40000	3.8103	0.39167	0.16501	125	-0.03	004.00	0
004	Fiber, Crude (%)	0413	3.8000	0.20000	3.8103	0.39167	0.16501	125	-0.03	004.07	0
004	Fiber, Crude (%)	0581	3.8000	0.20000	3.8103	0.39167	0.16501	125	-0.03	004.07	0
004	Fiber, Crude (%)	0629	3.8000	0.20000	3.8103	0.39167	0.16501	125	-0.03	004.99	0
004	Fiber, Crude (%)	0878	3.8250	0.15000	3.8103	0.39167	0.16501	125	0.04	004.06	0
004	Fiber, Crude (%)	0948	3.8400	0.20000	3.8103	0.39167	0.16501	125	0.08	004.06	0
004	Fiber, Crude (%)	0027	3.8455	0.09100	3.8103	0.39167	0.16501	125	0.09	004.06	0
004	Fiber, Crude (%)	0868	3.8550	0.03000	3.8103	0.39167	0.16501	125	0.11	004.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber, Crude (%)	2053	3.8750	0.11000	3.8103	0.39167	0.16501	125	0.17	004.07	0
004	Fiber, Crude (%)	2196	3.8800	0.00000	3.8103	0.39167	0.16501	125	0.18	004.07	0
004	Fiber, Crude (%)	0003	3.8850	0.57000	3.8103	0.39167	0.16501	125	0.19	004.07	0
004	Fiber, Crude (%)	0004	3.8900	0.04000	3.8103	0.39167	0.16501	125	0.20	004.07	0
004	Fiber, Crude (%)	0563	3.8950	0.05010	3.8103	0.39167	0.16501	125	0.22	004.07	0
004	Fiber, Crude (%)	0848	3.9000	0.00000	3.8103	0.39167	0.16501	125	0.23	004.07	0
004	Fiber, Crude (%)	2190	3.9050	0.09000	3.8103	0.39167	0.16501	125	0.24	004.07	0
004	Fiber, Crude (%)	0881	3.9060	0.13800	3.8103	0.39167	0.16501	125	0.24	004.06	0
004	Fiber, Crude (%)	0849	3.9215	0.08300	3.8103	0.39167	0.16501	125	0.28	004.07	0
004	Fiber, Crude (%)	0675	3.9300	0.12000	3.8103	0.39167	0.16501	125	0.31	004.06	0
004	Fiber, Crude (%)	0870	3.9312	0.05560	3.8103	0.39167	0.16501	125	0.31	004.07	0
004	Fiber, Crude (%)	0895	3.9350	0.01000	3.8103	0.39167	0.16501	125	0.32	004.06	0
004	Fiber, Crude (%)	0941	3.9500	0.26000	3.8103	0.39167	0.16501	125	0.36	004.06	0
004	Fiber, Crude (%)	0294	3.9500	0.50000	3.8103	0.39167	0.16501	125	0.36	004.07	0
004	Fiber, Crude (%)	0644	3.9895	0.13900	3.8103	0.39167	0.16501	125	0.46	004.07	0
004	Fiber, Crude (%)	0897	3.9950	0.01000	3.8103	0.39167	0.16501	125	0.47	004.06	0
004	Fiber, Crude (%)	0208	4.0000	0.24000	3.8103	0.39167	0.16501	125	0.48	004.00	0
004	Fiber, Crude (%)	0948	4.0100	0.18000	3.8103	0.39167	0.16501	125	0.51	004.07	0
004	Fiber, Crude (%)	2009	4.0389	0.07090	3.8103	0.39167	0.16501	125	0.58	004.07	0
004	Fiber, Crude (%)	0089	4.0400	0.00000	3.8103	0.39167	0.16501	125	0.59	004.07	0
004	Fiber, Crude (%)	2121	4.0400	0.02000	3.8103	0.39167	0.16501	125	0.59	004.11	0
004	Fiber, Crude (%)	0083	4.0450	0.39000	3.8103	0.39167	0.16501	125	0.60	004.07	0
004	Fiber, Crude (%)	0689	4.0500	0.10000	3.8103	0.39167	0.16501	125	0.61	004.06	0
004	Fiber, Crude (%)	0520	4.0500	0.10000	3.8103	0.39167	0.16501	125	0.61	004.07	0
004	Fiber, Crude (%)	0722	4.0645	0.05130	3.8103	0.39167	0.16501	125	0.65	004.06	0
004	Fiber, Crude (%)	0676	4.0700	0.12000	3.8103	0.39167	0.16501	125	0.66	004.06	0
004	Fiber, Crude (%)	0263	4.0816	0.07000	3.8103	0.39167	0.16501	125	0.69	004.07	0
004	Fiber, Crude (%)	2190	4.0950	0.23000	3.8103	0.39167	0.16501	125	0.73	004.03	0
004	Fiber, Crude (%)	0720	4.1000	0.00000	3.8103	0.39167	0.16501	125	0.74	004.11	0
004	Fiber, Crude (%)	0229	4.1100	0.16000	3.8103	0.39167	0.16501	125	0.77	004.07	0
004	Fiber, Crude (%)	0692	4.1500	0.30000	3.8103	0.39167	0.16501	125	0.87	004.07	0
004	Fiber, Crude (%)	0045	4.1850	0.09000	3.8103	0.39167	0.16501	125	0.96	004.03	0
004	Fiber, Crude (%)	0265	4.1850	0.07000	3.8103	0.39167	0.16501	125	0.96	004.07	0
004	Fiber, Crude (%)	2181	4.1850	0.05000	3.8103	0.39167	0.16501	125	0.96	004.07	0
004	Fiber, Crude (%)	0226	4.2000	0.00000	3.8103	0.39167	0.16501	125	1.00	004.00	0
004	Fiber, Crude (%)	2120	4.2400	0.08000	3.8103	0.39167	0.16501	125	1.10	004.11	0
004	Fiber, Crude (%)	2128	4.2576	0.21890	3.8103	0.39167	0.16501	125	1.14	004.11	0
004	Fiber, Crude (%)	0968	4.3025	0.06100	3.8103	0.39167	0.16501	125	1.26	004.06	0
004	Fiber, Crude (%)	0693	4.3500	0.10000	3.8103	0.39167	0.16501	125	1.38	004.03	0
004	Fiber, Crude (%)	2192	4.3650	0.05000	3.8103	0.39167	0.16501	125	1.42	004.03	0
004	Fiber, Crude (%)	0609	4.4500	0.04000	3.8103	0.39167	0.16501	125	1.63	004.06	0
004	Fiber, Crude (%)	0598	4.5650	0.27000	3.8103	0.39167	0.16501	125	1.93	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber, Crude (%)	0852	4.5700	0.04000	3.8103	0.39167	0.16501	125	1.94	004.11	0
004	Fiber, Crude (%)	0242	4.6450	0.07000	3.8103	0.39167	0.16501	125	2.13	004.07	0
004	Fiber, Crude (%)	0504	4.6600	0.44000	3.8103	0.39167	0.16501	125	2.17	004.00	0
004	Fiber, Crude (%)	0009	4.6650	0.37000	3.8103	0.39167	0.16501	125	2.18	004.07	0
004	Fiber, Crude (%)	0144	4.7850	0.13000	3.8103	0.39167	0.16501	125	2.49	004.07	0
004	Fiber, Crude (%)	0202	4.7900	0.10000	3.8103	0.39167	0.16501	125	2.50	004.07	0
004	Fiber, Crude (%)	0034	4.8400	0.56000	3.8103	0.39167	0.16501	125	2.63	004.07	0
004	Fiber, Crude (%)	0626	4.9750	0.55000	3.8103	0.39167	0.16501	125	2.97	004.07	0
004	Fiber, Crude (%)	0278	5.4500	0.30000	3.8103	0.39167	0.16501	125	4.19	004.07	0
004	Fiber, Crude (%)	0968	5.5250	0.23000	3.8103	0.39167	0.16501	125	4.38	004.07	0
004	Fiber, Crude (%)	0066	5.5900	0.16000	3.8103	0.39167	0.16501	125	4.54	004.07	0
004	Fiber, Crude (%)	2195	5.6150	0.25000	3.8103	0.39167	0.16501	125	4.61	004.07	0
004	Fiber, Crude (%)	0956	3.7000	0.80000	3.8103	0.39167	0.16501	125	-0.28	004.07	1
004	Fiber, Crude (%)	0553	4.6100	0.96000	3.8103	0.39167	0.16501	125	2.04	004.07	1
004	Fiber, Crude (%)	0889	0.00000	0.00000	3.8103	0.39167	0.16501	125		004.07	4
005	Ash (%)	0889	10.390	0.08000	14.932	0.57745	0.14447	143	-7.87	005.11	0
005	Ash (%)	0852	10.445	0.03000	14.932	0.57745	0.14447	143	-7.77	005.11	0
005	Ash (%)	0297	10.985	0.35000	14.932	0.57745	0.14447	143	-6.83	005.11	0
005	Ash (%)	0539	12.880	0.14000	14.932	0.57745	0.14447	143	-3.55	005.00	0
005	Ash (%)	0148	12.945	0.15000	14.932	0.57745	0.14447	143	-3.44	005.00	0
005	Ash (%)	0169	13.010	0.00000	14.932	0.57745	0.14447	143	-3.33	005.00	0
005	Ash (%)	2066	13.050	0.10000	14.932	0.57745	0.14447	143	-3.26	005.00	0
005	Ash (%)	0278	13.070	0.50000	14.932	0.57745	0.14447	143	-3.22	005.00	0
005	Ash (%)	0650	13.150	0.04000	14.932	0.57745	0.14447	143	-3.09	005.00	0
005	Ash (%)	0186	13.450	0.10000	14.932	0.57745	0.14447	143	-2.57	005.00	0
005	Ash (%)	0049	13.475	0.03000	14.932	0.57745	0.14447	143	-2.52	005.00	0
005	Ash (%)	0541	13.710	0.42000	14.932	0.57745	0.14447	143	-2.12	005.00	0
005	Ash (%)	0596	13.715	0.01000	14.932	0.57745	0.14447	143	-2.11	005.00	0
005	Ash (%)	0208	13.750	0.50000	14.932	0.57745	0.14447	143	-2.05	005.00	0
005	Ash (%)	0918	13.755	0.53000	14.932	0.57745	0.14447	143	-2.04	005.00	0
005	Ash (%)	0598	13.760	0.02000	14.932	0.57745	0.14447	143	-2.03	005.00	0
005	Ash (%)	0722	13.829	0.07410	14.932	0.57745	0.14447	143	-1.91	005.00	0
005	Ash (%)	0425	13.965	0.13000	14.932	0.57745	0.14447	143	-1.67	005.00	0
005	Ash (%)	0164	13.980	0.10000	14.932	0.57745	0.14447	143	-1.65	005.00	0
005	Ash (%)	0042	13.985	0.27000	14.932	0.57745	0.14447	143	-1.64	005.00	0
005	Ash (%)	0003	14.030	0.10000	14.932	0.57745	0.14447	143	-1.56	005.05	0
005	Ash (%)	0511	14.095	0.17000	14.932	0.57745	0.14447	143	-1.45	005.00	0
005	Ash (%)	2073	14.180	0.28000	14.932	0.57745	0.14447	143	-1.30	005.00	0
005	Ash (%)	0574	14.220	0.04000	14.932	0.57745	0.14447	143	-1.23	005.99	0
005	Ash (%)	2076	14.292	0.20250	14.932	0.57745	0.14447	143	-1.11	005.00	0
005	Ash (%)	0693	14.300	0.00000	14.932	0.57745	0.14447	143	-1.09	005.00	0
005	Ash (%)	2200	14.350	0.30000	14.932	0.57745	0.14447	143	-1.01	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0205	14.450	0.10000	14.932	0.57745	0.14447	143	-0.83	005.00	0
005	Ash (%)	0520	14.480	0.06000	14.932	0.57745	0.14447	143	-0.78	005.00	0
005	Ash (%)	0004	14.490	0.04000	14.932	0.57745	0.14447	143	-0.77	005.00	0
005	Ash (%)	0175	14.505	0.27000	14.932	0.57745	0.14447	143	-0.74	005.00	0
005	Ash (%)	0038	14.505	0.05000	14.932	0.57745	0.14447	143	-0.74	005.05	0
005	Ash (%)	0589	14.525	0.49000	14.932	0.57745	0.14447	143	-0.70	005.00	0
005	Ash (%)	2181	14.525	0.07000	14.932	0.57745	0.14447	143	-0.70	005.00	0
005	Ash (%)	2120	14.545	0.13000	14.932	0.57745	0.14447	143	-0.67	005.11	0
005	Ash (%)	0953	14.575	0.23000	14.932	0.57745	0.14447	143	-0.62	005.00	0
005	Ash (%)	0027	14.593	0.09500	14.932	0.57745	0.14447	143	-0.59	005.00	0
005	Ash (%)	0019	14.600	0.20000	14.932	0.57745	0.14447	143	-0.57	005.00	0
005	Ash (%)	0682	14.600	0.00000	14.932	0.57745	0.14447	143	-0.57	005.00	0
005	Ash (%)	0912	14.600	0.20000	14.932	0.57745	0.14447	143	-0.57	005.03	0
005	Ash (%)	0358	14.645	0.15000	14.932	0.57745	0.14447	143	-0.50	005.00	0
005	Ash (%)	0726	14.645	0.01000	14.932	0.57745	0.14447	143	-0.50	005.05	0
005	Ash (%)	0609	14.660	0.12000	14.932	0.57745	0.14447	143	-0.47	005.00	0
005	Ash (%)	0510	14.670	0.08000	14.932	0.57745	0.14447	143	-0.45	005.00	0
005	Ash (%)	0505	14.745	0.07000	14.932	0.57745	0.14447	143	-0.32	005.00	0
005	Ash (%)	0968	14.750	0.06000	14.932	0.57745	0.14447	143	-0.31	005.00	0
005	Ash (%)	2150	14.750	0.18000	14.932	0.57745	0.14447	143	-0.31	005.00	0
005	Ash (%)	0849	14.753	0.17400	14.932	0.57745	0.14447	143	-0.31	005.00	0
005	Ash (%)	0229	14.770	0.34000	14.932	0.57745	0.14447	143	-0.28	005.00	0
005	Ash (%)	0553	14.780	0.16000	14.932	0.57745	0.14447	143	-0.26	005.00	0
005	Ash (%)	2195	14.795	0.07000	14.932	0.57745	0.14447	143	-0.24	005.00	0
005	Ash (%)	0226	14.800	0.40000	14.932	0.57745	0.14447	143	-0.23	005.00	0
005	Ash (%)	0881	14.814	0.14000	14.932	0.57745	0.14447	143	-0.20	005.00	0
005	Ash (%)	2188	14.830	0.06000	14.932	0.57745	0.14447	143	-0.18	005.00	0
005	Ash (%)	2199	14.850	0.10000	14.932	0.57745	0.14447	143	-0.14	005.00	0
005	Ash (%)	0848	14.865	0.01000	14.932	0.57745	0.14447	143	-0.12	005.00	0
005	Ash (%)	0297	14.890	0.24000	14.932	0.57745	0.14447	143	-0.07	005.00	0
005	Ash (%)	0956	14.900	0.00000	14.932	0.57745	0.14447	143	-0.06	005.00	0
005	Ash (%)	0298	14.910	0.02000	14.932	0.57745	0.14447	143	-0.04	005.00	0
005	Ash (%)	0889	14.910	0.28000	14.932	0.57745	0.14447	143	-0.04	005.05	0
005	Ash (%)	0171	14.920	0.02000	14.932	0.57745	0.14447	143	-0.02	005.00	0
005	Ash (%)	0001	14.962	0.26340	14.932	0.57745	0.14447	143	0.05	005.00	0
005	Ash (%)	0142	14.970	0.08000	14.932	0.57745	0.14447	143	0.07	005.00	0
005	Ash (%)	0874	14.975	0.13000	14.932	0.57745	0.14447	143	0.07	005.00	0
005	Ash (%)	0062	14.982	0.02500	14.932	0.57745	0.14447	143	0.09	005.00	0
005	Ash (%)	0651	14.985	0.11100	14.932	0.57745	0.14447	143	0.09	005.00	0
005	Ash (%)	0035	14.985	0.07000	14.932	0.57745	0.14447	143	0.09	005.00	0
005	Ash (%)	0144	14.995	0.21000	14.932	0.57745	0.14447	143	0.11	005.00	0
005	Ash (%)	0354	15.000	0.08000	14.932	0.57745	0.14447	143	0.12	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0868	15.000	0.10000	14.932	0.57745	0.14447	143	0.12	005.05	0
005	Ash (%)	2006	15.010	0.22000	14.932	0.57745	0.14447	143	0.14	005.00	0
005	Ash (%)	0563	15.026	0.37370	14.932	0.57745	0.14447	143	0.16	005.00	0
005	Ash (%)	0675	15.030	0.40000	14.932	0.57745	0.14447	143	0.17	005.00	0
005	Ash (%)	0066	15.040	0.16000	14.932	0.57745	0.14447	143	0.19	005.05	0
005	Ash (%)	0629	15.040	0.04000	14.932	0.57745	0.14447	143	0.19	005.99	0
005	Ash (%)	0083	15.055	0.13000	14.932	0.57745	0.14447	143	0.21	005.00	0
005	Ash (%)	0108	15.060	0.14000	14.932	0.57745	0.14447	143	0.22	005.00	0
005	Ash (%)	0720	15.070	0.48000	14.932	0.57745	0.14447	143	0.24	005.05	0
005	Ash (%)	2121	15.070	0.20000	14.932	0.57745	0.14447	143	0.24	005.11	0
005	Ash (%)	0689	15.090	0.00000	14.932	0.57745	0.14447	143	0.27	005.05	0
005	Ash (%)	0870	15.096	0.39340	14.932	0.57745	0.14447	143	0.28	005.00	0
005	Ash (%)	0045	15.100	0.00000	14.932	0.57745	0.14447	143	0.29	005.00	0
005	Ash (%)	0033	15.150	0.30000	14.932	0.57745	0.14447	143	0.38	005.00	0
005	Ash (%)	0098	15.150	0.52000	14.932	0.57745	0.14447	143	0.38	005.00	0
005	Ash (%)	2193	15.155	0.07000	14.932	0.57745	0.14447	143	0.39	005.00	0
005	Ash (%)	0536	15.170	0.20000	14.932	0.57745	0.14447	143	0.41	005.99	0
005	Ash (%)	0263	15.185	0.00360	14.932	0.57745	0.14447	143	0.44	005.05	0
005	Ash (%)	2053	15.185	0.01000	14.932	0.57745	0.14447	143	0.44	005.00	0
005	Ash (%)	0644	15.190	0.06000	14.932	0.57745	0.14447	143	0.45	005.02	0
005	Ash (%)	0849	15.192	0.08300	14.932	0.57745	0.14447	143	0.45	005.05	0
005	Ash (%)	0047	15.210	0.20000	14.932	0.57745	0.14447	143	0.48	005.00	0
005	Ash (%)	0187	15.210	0.04000	14.932	0.57745	0.14447	143	0.48	005.00	0
005	Ash (%)	2196	15.210	0.00000	14.932	0.57745	0.14447	143	0.48	005.00	0
005	Ash (%)	0309	15.215	0.03500	14.932	0.57745	0.14447	143	0.49	005.00	0
005	Ash (%)	0885	15.219	0.10620	14.932	0.57745	0.14447	143	0.50	005.05	0
005	Ash (%)	0861	15.220	0.04000	14.932	0.57745	0.14447	143	0.50	005.00	0
005	Ash (%)	0265	15.250	0.10000	14.932	0.57745	0.14447	143	0.55	005.00	0
005	Ash (%)	0653	15.250	0.30000	14.932	0.57745	0.14447	143	0.55	005.00	0
005	Ash (%)	0948	15.255	0.01000	14.932	0.57745	0.14447	143	0.56	005.00	0
005	Ash (%)	0353	15.265	0.17000	14.932	0.57745	0.14447	143	0.58	005.00	0
005	Ash (%)	0940	15.265	0.51000	14.932	0.57745	0.14447	143	0.58	005.05	0
005	Ash (%)	0852	15.275	0.15000	14.932	0.57745	0.14447	143	0.59	005.00	0
005	Ash (%)	0242	15.300	0.10000	14.932	0.57745	0.14447	143	0.64	005.00	0
005	Ash (%)	0692	15.300	0.00000	14.932	0.57745	0.14447	143	0.64	005.05	0
005	Ash (%)	0504	15.310	0.30000	14.932	0.57745	0.14447	143	0.65	005.00	0
005	Ash (%)	0618	15.311	0.33280	14.932	0.57745	0.14447	143	0.66	005.00	0
005	Ash (%)	0878	15.315	0.19000	14.932	0.57745	0.14447	143	0.66	005.05	0
005	Ash (%)	0139	15.325	0.05000	14.932	0.57745	0.14447	143	0.68	005.00	0
005	Ash (%)	0676	15.325	0.37000	14.932	0.57745	0.14447	143	0.68	005.99	0
005	Ash (%)	2009	15.333	0.04350	14.932	0.57745	0.14447	143	0.69	005.05	0
005	Ash (%)	0202	15.345	0.03000	14.932	0.57745	0.14447	143	0.72	005.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0407	15.346	0.06110	14.932	0.57745	0.14447	143	0.72	005.00	0
005	Ash (%)	0294	15.350	0.10000	14.932	0.57745	0.14447	143	0.72	005.05	0
005	Ash (%)	0897	15.360	0.02000	14.932	0.57745	0.14447	143	0.74	005.05	0
005	Ash (%)	2109	15.365	0.13000	14.932	0.57745	0.14447	143	0.75	005.05	0
005	Ash (%)	0630	15.370	0.04000	14.932	0.57745	0.14447	143	0.76	005.00	0
005	Ash (%)	0941	15.370	0.20000	14.932	0.57745	0.14447	143	0.76	005.05	0
005	Ash (%)	2146	15.385	0.27000	14.932	0.57745	0.14447	143	0.78	005.05	0
005	Ash (%)	0622	15.395	0.30220	14.932	0.57745	0.14447	143	0.80	005.00	0
005	Ash (%)	0652	15.400	0.20000	14.932	0.57745	0.14447	143	0.81	005.99	0
005	Ash (%)	0100	15.460	0.04000	14.932	0.57745	0.14447	143	0.91	005.00	0
005	Ash (%)	0610	15.460	0.14000	14.932	0.57745	0.14447	143	0.91	005.05	0
005	Ash (%)	0590	15.470	0.06000	14.932	0.57745	0.14447	143	0.93	005.05	0
005	Ash (%)	0895	15.485	0.01000	14.932	0.57745	0.14447	143	0.96	005.05	0
005	Ash (%)	0939	15.495	0.17000	14.932	0.57745	0.14447	143	0.98	005.05	0
005	Ash (%)	0345	15.500	0.20000	14.932	0.57745	0.14447	143	0.98	005.00	0
005	Ash (%)	2161	15.500	0.20000	14.932	0.57745	0.14447	143	0.98	005.99	0
005	Ash (%)	2145	15.515	0.01000	14.932	0.57745	0.14447	143	1.01	005.05	0
005	Ash (%)	0683	15.525	0.05000	14.932	0.57745	0.14447	143	1.03	005.05	0
005	Ash (%)	2126	15.530	0.02000	14.932	0.57745	0.14447	143	1.04	005.05	0
005	Ash (%)	0152	15.550	0.06000	14.932	0.57745	0.14447	143	1.07	005.05	0
005	Ash (%)	0581	15.560	0.10000	14.932	0.57745	0.14447	143	1.09	005.00	0
005	Ash (%)	2192	15.565	0.05000	14.932	0.57745	0.14447	143	1.10	005.05	0
005	Ash (%)	0017	15.600	0.20000	14.932	0.57745	0.14447	143	1.16	005.00	0
005	Ash (%)	0884	15.600	0.20000	14.932	0.57745	0.14447	143	1.16	005.99	0
005	Ash (%)	2144	15.615	0.01000	14.932	0.57745	0.14447	143	1.18	005.99	0
005	Ash (%)	2190	15.700	0.02000	14.932	0.57745	0.14447	143	1.33	005.05	0
005	Ash (%)	0357	15.750	0.10000	14.932	0.57745	0.14447	143	1.42	005.00	0
005	Ash (%)	0910	15.750	0.10000	14.932	0.57745	0.14447	143	1.42	005.99	0
005	Ash (%)	0685	15.840	0.02000	14.932	0.57745	0.14447	143	1.57	005.05	0
005	Ash (%)	2089	15.855	0.19000	14.932	0.57745	0.14447	143	1.60	005.00	0
005	Ash (%)	0190	15.860	0.08000	14.932	0.57745	0.14447	143	1.61	005.05	0
005	Ash (%)	2129	16.435	0.09000	14.932	0.57745	0.14447	143	2.60	005.05	0
005	Ash (%)	0015	14.170	0.86000	14.932	0.57745	0.14447	143	-1.32	005.00	1
005	Ash (%)	0413	15.250	0.70000	14.932	0.57745	0.14447	143	0.55	005.05	1
005	Ash (%)	0720	5.9450	0.01000	14.932	0.57745	0.14447	143	-15.56	005.11	2
005	Ash (%)	2128	6.5577	0.02030	14.932	0.57745	0.14447	143	-14.50	005.11	2
006	Total Sugars (%)	0720	4.0850	0.05000	5.5657	0.75130	0.51058	6	-1.97	006.99	0
006	Total Sugars (%)	0910	5.2900	0.36000	5.5657	0.75130	0.51058	6	-0.37	006.00	0
006	Total Sugars (%)	0096	5.5857	1.3135	5.5657	0.75130	0.51058	6	0.03	006.99	0
006	Total Sugars (%)	2161	5.8100	0.12000	5.5657	0.75130	0.51058	6	0.33	006.01	0
006	Total Sugars (%)	0407	5.8700	0.22000	5.5657	0.75130	0.51058	6	0.40	006.01	0
006	Total Sugars (%)	0956	6.4000	1.0000	5.5657	0.75130	0.51058	6	1.11	006.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
008	Fiber, Acid Detergent (%)	0889	2.9350	0.11000	5.2305	0.61568	0.19403	58	-3.73	008.99	0
008	Fiber, Acid Detergent (%)	0676	3.9950	1.1700	5.2305	0.61568	0.19403	58	-2.01	008.02	0
008	Fiber, Acid Detergent (%)	0354	4.1250	0.07000	5.2305	0.61568	0.19403	58	-1.80	008.08	0
008	Fiber, Acid Detergent (%)	0956	4.1500	0.30000	5.2305	0.61568	0.19403	58	-1.76	008.08	0
008	Fiber, Acid Detergent (%)	0885	4.3019	0.00090	5.2305	0.61568	0.19403	58	-1.51	008.02	0
008	Fiber, Acid Detergent (%)	0910	4.3800	0.74000	5.2305	0.61568	0.19403	58	-1.38	008.08	0
008	Fiber, Acid Detergent (%)	0590	4.5000	0.20000	5.2305	0.61568	0.19403	58	-1.19	008.02	0
008	Fiber, Acid Detergent (%)	0848	4.5050	0.01000	5.2305	0.61568	0.19403	58	-1.18	008.08	0
008	Fiber, Acid Detergent (%)	0504	4.6300	0.30000	5.2305	0.61568	0.19403	58	-0.98	008.02	0
008	Fiber, Acid Detergent (%)	0413	4.7000	0.40000	5.2305	0.61568	0.19403	58	-0.86	008.08	0
008	Fiber, Acid Detergent (%)	0066	4.7300	0.16000	5.2305	0.61568	0.19403	58	-0.81	008.08	0
008	Fiber, Acid Detergent (%)	0208	4.7400	0.02000	5.2305	0.61568	0.19403	58	-0.80	008.02	0
008	Fiber, Acid Detergent (%)	0939	4.7850	0.25000	5.2305	0.61568	0.19403	58	-0.72	008.02	0
008	Fiber, Acid Detergent (%)	2146	4.8250	0.27000	5.2305	0.61568	0.19403	58	-0.66	008.08	0
008	Fiber, Acid Detergent (%)	0878	4.8300	0.20000	5.2305	0.61568	0.19403	58	-0.65	008.08	0
008	Fiber, Acid Detergent (%)	0861	4.8800	0.28000	5.2305	0.61568	0.19403	58	-0.57	008.08	0
008	Fiber, Acid Detergent (%)	2192	4.9600	0.48000	5.2305	0.61568	0.19403	58	-0.44	008.02	0
008	Fiber, Acid Detergent (%)	0013	4.9600	0.16000	5.2305	0.61568	0.19403	58	-0.44	008.08	0
008	Fiber, Acid Detergent (%)	0098	4.9950	0.07000	5.2305	0.61568	0.19403	58	-0.38	008.02	0
008	Fiber, Acid Detergent (%)	0610	5.0000	0.20000	5.2305	0.61568	0.19403	58	-0.37	008.99	0
008	Fiber, Acid Detergent (%)	0581	5.0200	0.10000	5.2305	0.61568	0.19403	58	-0.34	008.08	0
008	Fiber, Acid Detergent (%)	0948	5.0750	0.03000	5.2305	0.61568	0.19403	58	-0.25	008.02	0
008	Fiber, Acid Detergent (%)	0297	5.0900	0.06000	5.2305	0.61568	0.19403	58	-0.23	008.08	0
008	Fiber, Acid Detergent (%)	0187	5.1000	0.16000	5.2305	0.61568	0.19403	58	-0.21	008.02	0
008	Fiber, Acid Detergent (%)	2196	5.1000	0.00000	5.2305	0.61568	0.19403	58	-0.21	008.08	0
008	Fiber, Acid Detergent (%)	0309	5.1745	0.00040	5.2305	0.61568	0.19403	58	-0.09	008.02	0
008	Fiber, Acid Detergent (%)	0037	5.1750	0.33000	5.2305	0.61568	0.19403	58	-0.09	008.08	0
008	Fiber, Acid Detergent (%)	0870	5.1837	0.17520	5.2305	0.61568	0.19403	58	-0.08	008.08	0
008	Fiber, Acid Detergent (%)	0357	5.2000	0.20000	5.2305	0.61568	0.19403	58	-0.05	008.08	0
008	Fiber, Acid Detergent (%)	2190	5.2000	0.16000	5.2305	0.61568	0.19403	58	-0.05	008.08	0
008	Fiber, Acid Detergent (%)	0148	5.2300	0.16000	5.2305	0.61568	0.19403	58	0.00	008.08	0
008	Fiber, Acid Detergent (%)	0190	5.2450	0.27000	5.2305	0.61568	0.19403	58	0.02	008.08	0
008	Fiber, Acid Detergent (%)	0353	5.2550	0.39000	5.2305	0.61568	0.19403	58	0.04	008.02	0
008	Fiber, Acid Detergent (%)	0693	5.3000	0.00000	5.2305	0.61568	0.19403	58	0.11	008.08	0
008	Fiber, Acid Detergent (%)	0345	5.3400	0.10000	5.2305	0.61568	0.19403	58	0.18	008.02	0
008	Fiber, Acid Detergent (%)	0263	5.3400	0.06000	5.2305	0.61568	0.19403	58	0.18	008.08	0
008	Fiber, Acid Detergent (%)	0148	5.3450	0.07000	5.2305	0.61568	0.19403	58	0.19	008.02	0
008	Fiber, Acid Detergent (%)	0164	5.4000	0.00000	5.2305	0.61568	0.19403	58	0.28	008.08	0
008	Fiber, Acid Detergent (%)	0407	5.4207	0.31420	5.2305	0.61568	0.19403	58	0.31	008.08	0
008	Fiber, Acid Detergent (%)	0226	5.4500	0.10000	5.2305	0.61568	0.19403	58	0.36	008.02	0
008	Fiber, Acid Detergent (%)	0941	5.4800	0.30000	5.2305	0.61568	0.19403	58	0.41	008.99	0
008	Fiber, Acid Detergent (%)	0918	5.5144	0.61780	5.2305	0.61568	0.19403	58	0.46	008.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
008	Fiber, Acid Detergent (%)	0265	5.6200	0.14000	5.2305	0.61568	0.19403	58	0.63	008.05	0
008	Fiber, Acid Detergent (%)	0720	5.6800	0.02000	5.2305	0.61568	0.19403	58	0.73	008.99	0
008	Fiber, Acid Detergent (%)	0726	5.7050	0.05000	5.2305	0.61568	0.19403	58	0.77	008.08	0
008	Fiber, Acid Detergent (%)	0001	5.7300	0.16000	5.2305	0.61568	0.19403	58	0.81	008.08	0
008	Fiber, Acid Detergent (%)	2066	5.7550	0.11000	5.2305	0.61568	0.19403	58	0.85	008.02	0
008	Fiber, Acid Detergent (%)	0964	5.7715	0.04500	5.2305	0.61568	0.19403	58	0.88	008.02	0
008	Fiber, Acid Detergent (%)	0083	5.8350	0.07000	5.2305	0.61568	0.19403	58	0.98	008.08	0
008	Fiber, Acid Detergent (%)	0045	5.8400	0.08000	5.2305	0.61568	0.19403	58	0.99	008.08	0
008	Fiber, Acid Detergent (%)	2129	5.9550	0.19000	5.2305	0.61568	0.19403	58	1.18	008.08	0
008	Fiber, Acid Detergent (%)	0278	6.0500	0.10000	5.2305	0.61568	0.19403	58	1.33	008.08	0
008	Fiber, Acid Detergent (%)	0889	6.0650	0.19000	5.2305	0.61568	0.19403	58	1.36	008.08	0
008	Fiber, Acid Detergent (%)	0675	6.5650	0.01000	5.2305	0.61568	0.19403	58	2.17	008.02	0
008	Fiber, Acid Detergent (%)	0358	6.7800	0.66000	5.2305	0.61568	0.19403	58	2.52	008.08	0
008	Fiber, Acid Detergent (%)	0510	7.2000	0.20000	5.2305	0.61568	0.19403	58	3.20	008.08	0
008	Fiber, Acid Detergent (%)	0294	7.5500	0.10000	5.2305	0.61568	0.19403	58	3.77	008.08	0
008	Fiber, Acid Detergent (%)	0609	7.6800	0.14000	5.2305	0.61568	0.19403	58	3.98	008.02	0
008	Fiber, Acid Detergent (%)	0049	6.3750	2.7500	5.2305	0.61568	0.19403	58	1.86	008.08	1
008	Fiber, Acid Detergent (%)	0536	13.455	0.39000	5.2305	0.61568	0.19403	58	13.36	008.08	2
009	Fiber, Neutral Detergent (%)	0889	6.7950	0.19000	10.542	0.97806	0.31942	58	-3.83	009.99	0
009	Fiber, Neutral Detergent (%)	0910	9.0000	0.40000	10.542	0.97806	0.31942	58	-1.58	009.07	0
009	Fiber, Neutral Detergent (%)	0861	9.0250	0.39000	10.542	0.97806	0.31942	58	-1.55	009.09	0
009	Fiber, Neutral Detergent (%)	2146	9.3100	1.1400	10.542	0.97806	0.31942	58	-1.26	009.09	0
009	Fiber, Neutral Detergent (%)	0504	9.3300	0.40000	10.542	0.97806	0.31942	58	-1.24	009.04	0
009	Fiber, Neutral Detergent (%)	0049	9.5250	0.57000	10.542	0.97806	0.31942	58	-1.04	009.09	0
009	Fiber, Neutral Detergent (%)	0148	9.5400	0.04000	10.542	0.97806	0.31942	58	-1.02	009.09	0
009	Fiber, Neutral Detergent (%)	0294	9.5500	0.30000	10.542	0.97806	0.31942	58	-1.01	009.09	0
009	Fiber, Neutral Detergent (%)	0263	9.6300	0.00000	10.542	0.97806	0.31942	58	-0.93	009.09	0
009	Fiber, Neutral Detergent (%)	0956	9.7000	0.40000	10.542	0.97806	0.31942	58	-0.86	009.09	0
009	Fiber, Neutral Detergent (%)	0598	9.7200	0.20000	10.542	0.97806	0.31942	58	-0.84	009.09	0
009	Fiber, Neutral Detergent (%)	0187	9.7400	0.06000	10.542	0.97806	0.31942	58	-0.82	009.07	0
009	Fiber, Neutral Detergent (%)	0581	9.7550	0.11000	10.542	0.97806	0.31942	58	-0.80	009.09	0
009	Fiber, Neutral Detergent (%)	0358	9.7800	0.76000	10.542	0.97806	0.31942	58	-0.78	009.09	0
009	Fiber, Neutral Detergent (%)	0510	9.8000	0.20000	10.542	0.97806	0.31942	58	-0.76	009.09	0
009	Fiber, Neutral Detergent (%)	0407	9.8466	0.11090	10.542	0.97806	0.31942	58	-0.71	009.09	0
009	Fiber, Neutral Detergent (%)	0939	9.8750	0.13000	10.542	0.97806	0.31942	58	-0.68	009.09	0
009	Fiber, Neutral Detergent (%)	0353	9.8900	0.52000	10.542	0.97806	0.31942	58	-0.67	009.07	0
009	Fiber, Neutral Detergent (%)	0413	9.9500	0.30000	10.542	0.97806	0.31942	58	-0.60	009.09	0
009	Fiber, Neutral Detergent (%)	0001	9.9800	0.76000	10.542	0.97806	0.31942	58	-0.57	009.09	0
009	Fiber, Neutral Detergent (%)	0066	10.080	0.18000	10.542	0.97806	0.31942	58	-0.47	009.09	0
009	Fiber, Neutral Detergent (%)	0878	10.130	0.12000	10.542	0.97806	0.31942	58	-0.42	009.09	0
009	Fiber, Neutral Detergent (%)	2196	10.200	0.00000	10.542	0.97806	0.31942	58	-0.35	009.09	0
009	Fiber, Neutral Detergent (%)	0354	10.205	0.05000	10.542	0.97806	0.31942	58	-0.34	009.09	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
009	Fiber, Neutral Detergent (%)	2190	10.220	0.14000	10.542	0.97806	0.31942	58	-0.33	009.09	0
009	Fiber, Neutral Detergent (%)	0870	10.280	0.63420	10.542	0.97806	0.31942	58	-0.27	009.09	0
009	Fiber, Neutral Detergent (%)	2066	10.300	0.20000	10.542	0.97806	0.31942	58	-0.25	009.07	0
009	Fiber, Neutral Detergent (%)	2129	10.465	0.19000	10.542	0.97806	0.31942	58	-0.08	009.09	0
009	Fiber, Neutral Detergent (%)	0037	10.470	0.54000	10.542	0.97806	0.31942	58	-0.07	009.09	0
009	Fiber, Neutral Detergent (%)	0693	10.550	0.90000	10.542	0.97806	0.31942	58	0.01	009.09	0
009	Fiber, Neutral Detergent (%)	0676	10.645	0.53000	10.542	0.97806	0.31942	58	0.11	009.07	0
009	Fiber, Neutral Detergent (%)	0357	10.650	0.10000	10.542	0.97806	0.31942	58	0.11	009.09	0
009	Fiber, Neutral Detergent (%)	0675	10.720	0.02000	10.542	0.97806	0.31942	58	0.18	009.07	0
009	Fiber, Neutral Detergent (%)	0297	10.750	0.06000	10.542	0.97806	0.31942	58	0.21	009.09	0
009	Fiber, Neutral Detergent (%)	0098	10.800	0.80000	10.542	0.97806	0.31942	58	0.26	009.07	0
009	Fiber, Neutral Detergent (%)	0083	10.830	0.54000	10.542	0.97806	0.31942	58	0.29	009.09	0
009	Fiber, Neutral Detergent (%)	0885	10.833	0.18920	10.542	0.97806	0.31942	58	0.30	009.07	0
009	Fiber, Neutral Detergent (%)	0720	10.835	0.11000	10.542	0.97806	0.31942	58	0.30	009.99	0
009	Fiber, Neutral Detergent (%)	0918	10.837	0.62500	10.542	0.97806	0.31942	58	0.30	009.09	0
009	Fiber, Neutral Detergent (%)	0948	10.845	0.03000	10.542	0.97806	0.31942	58	0.31	009.09	0
009	Fiber, Neutral Detergent (%)	0265	10.950	0.44000	10.542	0.97806	0.31942	58	0.42	009.09	0
009	Fiber, Neutral Detergent (%)	0045	11.050	0.30000	10.542	0.97806	0.31942	58	0.52	009.09	0
009	Fiber, Neutral Detergent (%)	0278	11.050	0.50000	10.542	0.97806	0.31942	58	0.52	009.09	0
009	Fiber, Neutral Detergent (%)	0941	11.340	0.20000	10.542	0.97806	0.31942	58	0.82	009.99	0
009	Fiber, Neutral Detergent (%)	2192	11.395	1.1500	10.542	0.97806	0.31942	58	0.87	009.07	0
009	Fiber, Neutral Detergent (%)	0590	11.450	0.10000	10.542	0.97806	0.31942	58	0.93	009.07	0
009	Fiber, Neutral Detergent (%)	0226	11.500	0.20000	10.542	0.97806	0.31942	58	0.98	009.07	0
009	Fiber, Neutral Detergent (%)	0848	11.500	0.00000	10.542	0.97806	0.31942	58	0.98	009.09	0
009	Fiber, Neutral Detergent (%)	0345	11.525	0.15000	10.542	0.97806	0.31942	58	1.01	009.07	0
009	Fiber, Neutral Detergent (%)	0726	11.540	0.06000	10.542	0.97806	0.31942	58	1.02	009.09	0
009	Fiber, Neutral Detergent (%)	0190	11.730	0.14000	10.542	0.97806	0.31942	58	1.22	009.09	0
009	Fiber, Neutral Detergent (%)	0208	11.800	0.00000	10.542	0.97806	0.31942	58	1.29	009.07	0
009	Fiber, Neutral Detergent (%)	0610	11.850	0.50000	10.542	0.97806	0.31942	58	1.34	009.07	0
009	Fiber, Neutral Detergent (%)	0309	11.910	0.01710	10.542	0.97806	0.31942	58	1.40	009.07	0
009	Fiber, Neutral Detergent (%)	0889	12.405	0.67000	10.542	0.97806	0.31942	58	1.91	009.09	0
009	Fiber, Neutral Detergent (%)	0164	12.600	0.40000	10.542	0.97806	0.31942	58	2.10	009.09	0
009	Fiber, Neutral Detergent (%)	0609	13.460	0.22000	10.542	0.97806	0.31942	58	2.98	009.07	0
009	Fiber, Neutral Detergent (%)	0536	14.620	0.54000	10.542	0.97806	0.31942	58	4.17	009.09	0
010	Moisture (%)	2161	6.9050	0.05000	8.5107	0.51972	0.09186	29	-3.09	010.99	0
010	Moisture (%)	0843	7.6300	0.12000	8.5107	0.51972	0.09186	29	-1.69	010.03	0
010	Moisture (%)	0028	7.7950	0.35000	8.5107	0.51972	0.09186	29	-1.38	010.99	0
010	Moisture (%)	2129	7.9800	0.26000	8.5107	0.51972	0.09186	29	-1.02	010.99	0
010	Moisture (%)	0722	8.0203	0.09690	8.5107	0.51972	0.09186	29	-0.94	010.99	0
010	Moisture (%)	0964	8.0365	0.26900	8.5107	0.51972	0.09186	29	-0.91	010.99	0
010	Moisture (%)	0948	8.1350	0.05000	8.5107	0.51972	0.09186	29	-0.72	010.99	0
010	Moisture (%)	2144	8.2200	0.02000	8.5107	0.51972	0.09186	29	-0.56	010.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
010	Moisture (%)	0939	8.2600	0.00000	8.5107	0.51972	0.09186	29	-0.48	010.99	0
010	Moisture (%)	0917	8.3000	0.00000	8.5107	0.51972	0.09186	29	-0.41	010.99	0
010	Moisture (%)	0884	8.3500	0.10000	8.5107	0.51972	0.09186	29	-0.31	010.99	0
010	Moisture (%)	0164	8.4150	0.09000	8.5107	0.51972	0.09186	29	-0.18	010.03	0
010	Moisture (%)	2190	8.4300	0.12000	8.5107	0.51972	0.09186	29	-0.16	010.11	0
010	Moisture (%)	2120	8.4600	0.02000	8.5107	0.51972	0.09186	29	-0.10	010.11	0
010	Moisture (%)	2178	8.5200	0.20000	8.5107	0.51972	0.09186	29	0.02	010.99	0
010	Moisture (%)	0720	8.5350	0.01000	8.5107	0.51972	0.09186	29	0.05	010.11	0
010	Moisture (%)	0405	8.6500	0.08000	8.5107	0.51972	0.09186	29	0.27	010.99	0
010	Moisture (%)	0652	8.6500	0.10000	8.5107	0.51972	0.09186	29	0.27	010.99	0
010	Moisture (%)	2121	8.6750	0.05000	8.5107	0.51972	0.09186	29	0.32	010.11	0
010	Moisture (%)	0190	8.6850	0.01000	8.5107	0.51972	0.09186	29	0.34	010.99	0
010	Moisture (%)	2076	8.7050	0.07000	8.5107	0.51972	0.09186	29	0.37	010.99	0
010	Moisture (%)	2089	8.7100	0.12000	8.5107	0.51972	0.09186	29	0.38	010.99	0
010	Moisture (%)	0726	8.7750	0.03000	8.5107	0.51972	0.09186	29	0.51	010.99	0
010	Moisture (%)	0889	8.9150	0.01000	8.5107	0.51972	0.09186	29	0.78	010.11	0
010	Moisture (%)	0152	8.9900	0.06000	8.5107	0.51972	0.09186	29	0.92	010.11	0
010	Moisture (%)	0952	9.2650	0.03000	8.5107	0.51972	0.09186	29	1.45	010.99	0
010	Moisture (%)	2127	9.4900	0.00000	8.5107	0.51972	0.09186	29	1.88	010.99	0
010	Moisture (%)	2128	10.176	0.26800	8.5107	0.51972	0.09186	29	3.20	010.11	0
010	Moisture (%)	0852	10.360	0.08000	8.5107	0.51972	0.09186	29	3.56	010.11	0
010	Moisture (%)	0692	8.3400	0.50000	8.5107	0.51972	0.09186	29	-0.33	010.99	1
011	Loss on Drying (%)	0536	7.2850	0.01000	9.1117	0.32085	0.09423	68	-5.69	011.01	0
011	Loss on Drying (%)	0345	7.9950	0.13000	9.1117	0.32085	0.09423	68	-3.48	011.01	0
011	Loss on Drying (%)	2200	8.0000	0.20000	9.1117	0.32085	0.09423	68	-3.46	011.01	0
011	Loss on Drying (%)	0265	8.3100	0.08000	9.1117	0.32085	0.09423	68	-2.50	011.01	0
011	Loss on Drying (%)	2073	8.4150	0.05000	9.1117	0.32085	0.09423	68	-2.17	011.01	0
011	Loss on Drying (%)	2097	8.4174	0.08760	9.1117	0.32085	0.09423	68	-2.16	011.01	0
011	Loss on Drying (%)	0874	8.4700	0.14000	9.1117	0.32085	0.09423	68	-2.00	011.01	0
011	Loss on Drying (%)	2193	8.5050	0.15000	9.1117	0.32085	0.09423	68	-1.89	011.01	0
011	Loss on Drying (%)	0574	8.6350	0.03000	9.1117	0.32085	0.09423	68	-1.49	011.01	0
011	Loss on Drying (%)	0563	8.7370	0.06000	9.1117	0.32085	0.09423	68	-1.17	011.01	0
011	Loss on Drying (%)	2022	8.8000	0.00000	9.1117	0.32085	0.09423	68	-0.97	011.01	0
011	Loss on Drying (%)	0685	8.8350	0.03000	9.1117	0.32085	0.09423	68	-0.86	011.01	0
011	Loss on Drying (%)	0859	8.8355	0.10300	9.1117	0.32085	0.09423	68	-0.86	011.01	0
011	Loss on Drying (%)	0298	8.8500	0.02000	9.1117	0.32085	0.09423	68	-0.82	011.01	0
011	Loss on Drying (%)	2181	8.9050	0.07000	9.1117	0.32085	0.09423	68	-0.64	011.01	0
011	Loss on Drying (%)	2188	8.9150	0.03000	9.1117	0.32085	0.09423	68	-0.61	011.01	0
011	Loss on Drying (%)	0942	8.9150	0.07000	9.1117	0.32085	0.09423	68	-0.61	011.02	0
011	Loss on Drying (%)	0675	8.9350	0.07000	9.1117	0.32085	0.09423	68	-0.55	011.01	0
011	Loss on Drying (%)	0108	8.9550	0.03000	9.1117	0.32085	0.09423	68	-0.49	011.01	0
011	Loss on Drying (%)	0242	8.9550	0.03000	9.1117	0.32085	0.09423	68	-0.49	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
011	Loss on Drying (%)	0881	8.9930	0.18800	9.1117	0.32085	0.09423	68	-0.37	011.01	0
011	Loss on Drying (%)	0175	9.0000	0.40000	9.1117	0.32085	0.09423	68	-0.35	011.01	0
011	Loss on Drying (%)	0062	9.0270	0.05000	9.1117	0.32085	0.09423	68	-0.26	011.01	0
011	Loss on Drying (%)	0202	9.0400	0.04000	9.1117	0.32085	0.09423	68	-0.22	011.01	0
011	Loss on Drying (%)	0208	9.0450	0.03000	9.1117	0.32085	0.09423	68	-0.21	011.01	0
011	Loss on Drying (%)	0857	9.0500	0.10000	9.1117	0.32085	0.09423	68	-0.19	011.99	0
011	Loss on Drying (%)	0653	9.0550	0.03000	9.1117	0.32085	0.09423	68	-0.18	011.01	0
011	Loss on Drying (%)	0953	9.0650	0.01000	9.1117	0.32085	0.09423	68	-0.15	011.01	0
011	Loss on Drying (%)	0598	9.0700	0.22000	9.1117	0.32085	0.09423	68	-0.13	011.01	0
011	Loss on Drying (%)	0148	9.0800	0.32000	9.1117	0.32085	0.09423	68	-0.10	011.01	0
011	Loss on Drying (%)	0171	9.0850	0.03000	9.1117	0.32085	0.09423	68	-0.08	011.01	0
011	Loss on Drying (%)	0968	9.0860	0.00400	9.1117	0.32085	0.09423	68	-0.08	011.01	0
011	Loss on Drying (%)	2168	9.1000	0.02000	9.1117	0.32085	0.09423	68	-0.04	011.01	0
011	Loss on Drying (%)	2196	9.1000	0.00000	9.1117	0.32085	0.09423	68	-0.04	011.01	0
011	Loss on Drying (%)	0425	9.1250	0.05000	9.1117	0.32085	0.09423	68	0.04	011.02	0
011	Loss on Drying (%)	0407	9.1434	0.09730	9.1117	0.32085	0.09423	68	0.10	011.01	0
011	Loss on Drying (%)	0520	9.1500	0.10000	9.1117	0.32085	0.09423	68	0.12	011.01	0
011	Loss on Drying (%)	0164	9.1500	0.16000	9.1117	0.32085	0.09423	68	0.12	011.01	0
011	Loss on Drying (%)	2006	9.1600	0.16000	9.1117	0.32085	0.09423	68	0.15	011.01	0
011	Loss on Drying (%)	0651	9.1865	0.04300	9.1117	0.32085	0.09423	68	0.23	011.01	0
011	Loss on Drying (%)	0229	9.1865	0.00900	9.1117	0.32085	0.09423	68	0.23	011.01	0
011	Loss on Drying (%)	0294	9.2000	0.00000	9.1117	0.32085	0.09423	68	0.28	011.01	0
011	Loss on Drying (%)	0682	9.2000	0.00000	9.1117	0.32085	0.09423	68	0.28	011.01	0
011	Loss on Drying (%)	0722	9.2047	0.05960	9.1117	0.32085	0.09423	68	0.29	011.01	0
011	Loss on Drying (%)	0870	9.2089	0.06580	9.1117	0.32085	0.09423	68	0.30	011.01	0
011	Loss on Drying (%)	0539	9.2300	0.02000	9.1117	0.32085	0.09423	68	0.37	011.01	0
011	Loss on Drying (%)	0226	9.2400	0.12000	9.1117	0.32085	0.09423	68	0.40	011.01	0
011	Loss on Drying (%)	0033	9.2500	0.00000	9.1117	0.32085	0.09423	68	0.43	011.01	0
011	Loss on Drying (%)	0144	9.2550	0.15000	9.1117	0.32085	0.09423	68	0.45	011.01	0
011	Loss on Drying (%)	0354	9.2600	0.02000	9.1117	0.32085	0.09423	68	0.46	011.01	0
011	Loss on Drying (%)	0100	9.2750	0.03000	9.1117	0.32085	0.09423	68	0.51	011.01	0
011	Loss on Drying (%)	0511	9.2750	0.13000	9.1117	0.32085	0.09423	68	0.51	011.01	0
011	Loss on Drying (%)	0098	9.3250	0.11000	9.1117	0.32085	0.09423	68	0.66	011.01	0
011	Loss on Drying (%)	0510	9.3500	0.10000	9.1117	0.32085	0.09423	68	0.74	011.01	0
011	Loss on Drying (%)	0047	9.3800	0.54000	9.1117	0.32085	0.09423	68	0.84	011.01	0
011	Loss on Drying (%)	0066	9.3900	0.08000	9.1117	0.32085	0.09423	68	0.87	011.01	0
011	Loss on Drying (%)	0861	9.4350	0.03000	9.1117	0.32085	0.09423	68	1.01	011.01	0
011	Loss on Drying (%)	0541	9.4500	0.16000	9.1117	0.32085	0.09423	68	1.05	011.01	0
011	Loss on Drying (%)	0309	9.4650	0.23000	9.1117	0.32085	0.09423	68	1.10	011.01	0
011	Loss on Drying (%)	0843	9.4900	0.22000	9.1117	0.32085	0.09423	68	1.18	011.01	0
011	Loss on Drying (%)	0263	9.5200	0.04000	9.1117	0.32085	0.09423	68	1.27	011.01	0
011	Loss on Drying (%)	0650	9.5300	0.22000	9.1117	0.32085	0.09423	68	1.30	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
011	Loss on Drying (%)	0848	9.5300	0.06000	9.1117	0.32085	0.09423	68	1.30	011.01	0
011	Loss on Drying (%)	0205	9.5650	0.03000	9.1117	0.32085	0.09423	68	1.41	011.01	0
011	Loss on Drying (%)	0589	9.5750	0.13000	9.1117	0.32085	0.09423	68	1.44	011.01	0
011	Loss on Drying (%)	0358	9.6150	0.03000	9.1117	0.32085	0.09423	68	1.57	011.01	0
011	Loss on Drying (%)	0004	9.7400	0.06000	9.1117	0.32085	0.09423	68	1.96	011.01	0
011	Loss on Drying (%)	0852	10.590	0.32000	9.1117	0.32085	0.09423	68	4.61	011.99	0
011	Loss on Drying (%)	0553	8.8950	1.2100	9.1117	0.32085	0.09423	68	-0.68	011.01	1
012	Starch (%)	0918	22.337	1.6460	30.394	2.2355	0.40545	43	-3.60	012.01	0
012	Starch (%)	0609	23.515	0.37000	30.394	2.2355	0.40545	43	-3.08	012.00	0
012	Starch (%)	0164	27.400	0.80000	30.394	2.2355	0.40545	43	-1.34	012.01	0
012	Starch (%)	0510	27.400	0.20000	30.394	2.2355	0.40545	43	-1.34	012.04	0
012	Starch (%)	0889	27.590	0.04000	30.394	2.2355	0.40545	43	-1.25	012.11	0
012	Starch (%)	0676	27.895	0.05000	30.394	2.2355	0.40545	43	-1.12	012.01	0
012	Starch (%)	0297	27.900	0.06000	30.394	2.2355	0.40545	43	-1.12	012.03	0
012	Starch (%)	2195	28.020	0.02000	30.394	2.2355	0.40545	43	-1.06	012.00	0
012	Starch (%)	0956	28.150	0.70000	30.394	2.2355	0.40545	43	-1.00	012.01	0
012	Starch (%)	0407	28.270	1.0200	30.394	2.2355	0.40545	43	-0.95	012.03	0
012	Starch (%)	0160	28.860	0.26000	30.394	2.2355	0.40545	43	-0.69	012.01	0
012	Starch (%)	0278	29.250	1.1000	30.394	2.2355	0.40545	43	-0.51	012.04	0
012	Starch (%)	0098	29.400	0.60000	30.394	2.2355	0.40545	43	-0.44	012.03	0
012	Starch (%)	0045	29.450	1.1000	30.394	2.2355	0.40545	43	-0.42	012.01	0
012	Starch (%)	2193	29.515	0.21000	30.394	2.2355	0.40545	43	-0.39	012.00	0
012	Starch (%)	0870	29.881	0.11090	30.394	2.2355	0.40545	43	-0.23	012.01	0
012	Starch (%)	2188	29.940	0.10000	30.394	2.2355	0.40545	43	-0.20	012.00	0
012	Starch (%)	0689	30.000	0.20000	30.394	2.2355	0.40545	43	-0.18	012.00	0
012	Starch (%)	0353	30.040	1.0000	30.394	2.2355	0.40545	43	-0.16	012.04	0
012	Starch (%)	0955	30.250	0.30000	30.394	2.2355	0.40545	43	-0.06	012.03	0
012	Starch (%)	0948	30.455	0.19000	30.394	2.2355	0.40545	43	0.03	012.04	0
012	Starch (%)	0941	30.535	0.33000	30.394	2.2355	0.40545	43	0.06	012.00	0
012	Starch (%)	0208	30.600	0.00000	30.394	2.2355	0.40545	43	0.09	012.04	0
012	Starch (%)	0164	30.750	0.10000	30.394	2.2355	0.40545	43	0.16	012.00	0
012	Starch (%)	0889	30.900	0.24000	30.394	2.2355	0.40545	43	0.23	012.03	0
012	Starch (%)	0354	30.925	0.61000	30.394	2.2355	0.40545	43	0.24	012.00	0
012	Starch (%)	2196	31.180	0.00000	30.394	2.2355	0.40545	43	0.35	012.00	0
012	Starch (%)	2181	31.300	0.00000	30.394	2.2355	0.40545	43	0.41	012.00	0
012	Starch (%)	2009	31.357	0.16330	30.394	2.2355	0.40545	43	0.43	012.00	0
012	Starch (%)	2006	31.390	0.02000	30.394	2.2355	0.40545	43	0.45	012.00	0
012	Starch (%)	0683	31.415	0.11000	30.394	2.2355	0.40545	43	0.46	012.00	0
012	Starch (%)	0693	31.450	1.1000	30.394	2.2355	0.40545	43	0.47	012.01	0
012	Starch (%)	0878	31.750	0.10000	30.394	2.2355	0.40545	43	0.61	012.00	0
012	Starch (%)	0939	31.955	0.43000	30.394	2.2355	0.40545	43	0.70	012.00	0
012	Starch (%)	0910	32.050	0.10000	30.394	2.2355	0.40545	43	0.74	012.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
012	Starch (%)	0720	32.065	0.13000	30.394	2.2355	0.40545	43	0.75	012.11	0
012	Starch (%)	0610	32.400	0.20000	30.394	2.2355	0.40545	43	0.90	012.00	0
012	Starch (%)	0297	32.520	0.02000	30.394	2.2355	0.40545	43	0.95	012.11	0
012	Starch (%)	0265	33.655	1.8700	30.394	2.2355	0.40545	43	1.46	012.01	0
012	Starch (%)	0852	33.805	0.07000	30.394	2.2355	0.40545	43	1.53	012.11	0
012	Starch (%)	2089	33.900	0.46000	30.394	2.2355	0.40545	43	1.57	012.01	0
012	Starch (%)	2128	35.527	0.35400	30.394	2.2355	0.40545	43	2.30	012.11	0
012	Starch (%)	0848	35.935	0.95000	30.394	2.2355	0.40545	43	2.48	012.03	0
013	Fat, Acid Pretreat (%)	0309	3.5167	0.70360	4.4187	0.57301	0.17191	51	-1.57	013.00	0
013	Fat, Acid Pretreat (%)	2145	3.5600	0.04000	4.4187	0.57301	0.17191	51	-1.50	013.00	0
013	Fat, Acid Pretreat (%)	0968	3.6230	0.06000	4.4187	0.57301	0.17191	51	-1.39	013.13	0
013	Fat, Acid Pretreat (%)	0083	3.6700	0.24000	4.4187	0.57301	0.17191	51	-1.31	013.00	0
013	Fat, Acid Pretreat (%)	0407	3.7450	0.07000	4.4187	0.57301	0.17191	51	-1.18	013.13	0
013	Fat, Acid Pretreat (%)	0885	3.7925	0.07470	4.4187	0.57301	0.17191	51	-1.09	013.00	0
013	Fat, Acid Pretreat (%)	2128	3.8104	0.01880	4.4187	0.57301	0.17191	51	-1.06	013.12	0
013	Fat, Acid Pretreat (%)	0689	3.8500	0.10000	4.4187	0.57301	0.17191	51	-0.99	013.00	0
013	Fat, Acid Pretreat (%)	0036	3.8680	0.07270	4.4187	0.57301	0.17191	51	-0.96	013.10	0
013	Fat, Acid Pretreat (%)	2146	3.8800	0.32000	4.4187	0.57301	0.17191	51	-0.94	013.00	0
013	Fat, Acid Pretreat (%)	0297	3.9000	0.18000	4.4187	0.57301	0.17191	51	-0.91	013.00	0
013	Fat, Acid Pretreat (%)	0242	3.9400	0.08000	4.4187	0.57301	0.17191	51	-0.84	013.00	0
013	Fat, Acid Pretreat (%)	0504	3.9900	0.12000	4.4187	0.57301	0.17191	51	-0.75	013.00	0
013	Fat, Acid Pretreat (%)	0955	3.9950	0.17000	4.4187	0.57301	0.17191	51	-0.74	013.13	0
013	Fat, Acid Pretreat (%)	0948	4.0650	0.01000	4.4187	0.57301	0.17191	51	-0.62	013.00	0
013	Fat, Acid Pretreat (%)	0610	4.0950	0.03000	4.4187	0.57301	0.17191	51	-0.56	013.10	0
013	Fat, Acid Pretreat (%)	0618	4.1564	0.25500	4.4187	0.57301	0.17191	51	-0.46	013.08	0
013	Fat, Acid Pretreat (%)	0581	4.2000	0.20000	4.4187	0.57301	0.17191	51	-0.38	013.13	0
013	Fat, Acid Pretreat (%)	0353	4.2200	0.16000	4.4187	0.57301	0.17191	51	-0.35	013.10	0
013	Fat, Acid Pretreat (%)	0910	4.2700	0.36000	4.4187	0.57301	0.17191	51	-0.26	013.00	0
013	Fat, Acid Pretreat (%)	0884	4.3000	0.00000	4.4187	0.57301	0.17191	51	-0.21	013.00	0
013	Fat, Acid Pretreat (%)	0265	4.3050	0.45000	4.4187	0.57301	0.17191	51	-0.20	013.13	0
013	Fat, Acid Pretreat (%)	2006	4.3100	0.38000	4.4187	0.57301	0.17191	51	-0.19	013.10	0
013	Fat, Acid Pretreat (%)	0019	4.3150	0.13000	4.4187	0.57301	0.17191	51	-0.18	013.13	0
013	Fat, Acid Pretreat (%)	0652	4.3500	0.10000	4.4187	0.57301	0.17191	51	-0.12	013.00	0
013	Fat, Acid Pretreat (%)	0074	4.3700	0.34000	4.4187	0.57301	0.17191	51	-0.09	013.02	0
013	Fat, Acid Pretreat (%)	0098	4.3900	0.02000	4.4187	0.57301	0.17191	51	-0.05	013.00	0
013	Fat, Acid Pretreat (%)	0229	4.4700	0.02000	4.4187	0.57301	0.17191	51	0.09	013.00	0
013	Fat, Acid Pretreat (%)	0139	4.4800	0.00000	4.4187	0.57301	0.17191	51	0.11	013.00	0
013	Fat, Acid Pretreat (%)	0541	4.5150	0.03000	4.4187	0.57301	0.17191	51	0.17	013.00	0
013	Fat, Acid Pretreat (%)	0148	4.5250	0.13000	4.4187	0.57301	0.17191	51	0.19	013.02	0
013	Fat, Acid Pretreat (%)	0843	4.5300	0.44000	4.4187	0.57301	0.17191	51	0.19	013.13	0
013	Fat, Acid Pretreat (%)	0208	4.6050	0.31000	4.4187	0.57301	0.17191	51	0.33	013.02	0
013	Fat, Acid Pretreat (%)	0354	4.6200	0.18000	4.4187	0.57301	0.17191	51	0.35	013.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
013	Fat, Acid Pretreat (%)	0019	4.6500	0.24000	4.4187	0.57301	0.17191	51	0.40	013.02	0
013	Fat, Acid Pretreat (%)	0171	4.7250	0.13000	4.4187	0.57301	0.17191	51	0.53	013.02	0
013	Fat, Acid Pretreat (%)	0010	4.8000	0.00000	4.4187	0.57301	0.17191	51	0.67	013.02	0
013	Fat, Acid Pretreat (%)	0874	4.8100	0.16000	4.4187	0.57301	0.17191	51	0.68	013.02	0
013	Fat, Acid Pretreat (%)	0553	4.8500	0.18000	4.4187	0.57301	0.17191	51	0.75	013.02	0
013	Fat, Acid Pretreat (%)	0164	4.8600	0.02000	4.4187	0.57301	0.17191	51	0.77	013.02	0
013	Fat, Acid Pretreat (%)	0650	4.9100	0.48000	4.4187	0.57301	0.17191	51	0.86	013.02	0
013	Fat, Acid Pretreat (%)	2073	4.9400	0.46000	4.4187	0.57301	0.17191	51	0.91	013.00	0
013	Fat, Acid Pretreat (%)	0870	5.0111	0.23730	4.4187	0.57301	0.17191	51	1.03	013.02	0
013	Fat, Acid Pretreat (%)	0202	5.0400	0.00000	4.4187	0.57301	0.17191	51	1.08	013.00	0
013	Fat, Acid Pretreat (%)	0100	5.0600	0.32000	4.4187	0.57301	0.17191	51	1.12	013.02	0
013	Fat, Acid Pretreat (%)	2150	5.1550	0.07000	4.4187	0.57301	0.17191	51	1.28	013.02	0
013	Fat, Acid Pretreat (%)	0676	5.2200	0.18000	4.4187	0.57301	0.17191	51	1.40	013.02	0
013	Fat, Acid Pretreat (%)	0675	5.2450	0.27000	4.4187	0.57301	0.17191	51	1.44	013.02	0
013	Fat, Acid Pretreat (%)	2076	5.2490	0.16540	4.4187	0.57301	0.17191	51	1.45	013.00	0
013	Fat, Acid Pretreat (%)	0682	5.3200	0.00000	4.4187	0.57301	0.17191	51	1.57	013.02	0
013	Fat, Acid Pretreat (%)	0861	5.4600	0.06000	4.4187	0.57301	0.17191	51	1.82	013.02	0
015	Aluminum (mg / kg (ppm))	0049	169.20	7.3600	228.56	34.959	8.0467	13	-1.70	015.41	0
015	Aluminum (mg / kg (ppm))	0169	191.50	13.000	228.56	34.959	8.0467	13	-1.06	015.43	0
015	Aluminum (mg / kg (ppm))	0186	203.50	1.0000	228.56	34.959	8.0467	13	-0.72	015.52	0
015	Aluminum (mg / kg (ppm))	0918	205.54	16.248	228.56	34.959	8.0467	13	-0.66	015.43	0
015	Aluminum (mg / kg (ppm))	0345	214.50	5.0000	228.56	34.959	8.0467	13	-0.40	015.43	0
015	Aluminum (mg / kg (ppm))	0353	225.00	4.0000	228.56	34.959	8.0467	13	-0.10	015.43	0
015	Aluminum (mg / kg (ppm))	0510	225.00	6.0000	228.56	34.959	8.0467	13	-0.10	015.43	0
015	Aluminum (mg / kg (ppm))	0297	232.50	15.000	228.56	34.959	8.0467	13	0.11	015.43	0
015	Aluminum (mg / kg (ppm))	0560	234.60	16.800	228.56	34.959	8.0467	13	0.17	015.42	0
015	Aluminum (mg / kg (ppm))	0520	257.00	8.0000	228.56	34.959	8.0467	13	0.81	015.41	0
015	Aluminum (mg / kg (ppm))	0164	261.50	5.0000	228.56	34.959	8.0467	13	0.94	015.41	0
015	Aluminum (mg / kg (ppm))	0171	264.00	4.2000	228.56	34.959	8.0467	13	1.01	015.41	0
015	Aluminum (mg / kg (ppm))	0553	280.50	3.0000	228.56	34.959	8.0467	13	1.49	015.53	0
017	Boron (mg / kg (ppm))	0297	6.5000	1.0000	12.901	1.3882	0.52813	16	-4.61	017.43	0
017	Boron (mg / kg (ppm))	0226	11.245	0.05000	12.901	1.3882	0.52813	16	-1.19	017.41	0
017	Boron (mg / kg (ppm))	0265	11.950	1.1000	12.901	1.3882	0.52813	16	-0.69	017.42	0
017	Boron (mg / kg (ppm))	0345	11.950	0.50000	12.901	1.3882	0.52813	16	-0.69	017.43	0
017	Boron (mg / kg (ppm))	0358	12.430	1.0000	12.901	1.3882	0.52813	16	-0.34	017.41	0
017	Boron (mg / kg (ppm))	0032	12.450	0.10000	12.901	1.3882	0.52813	16	-0.33	017.53	0
017	Boron (mg / kg (ppm))	0045	12.500	1.0000	12.901	1.3882	0.52813	16	-0.29	017.42	0
017	Boron (mg / kg (ppm))	0353	12.500	1.0000	12.901	1.3882	0.52813	16	-0.29	017.43	0
017	Boron (mg / kg (ppm))	0049	12.785	0.51000	12.901	1.3882	0.52813	16	-0.08	017.41	0
017	Boron (mg / kg (ppm))	0693	13.000	0.00000	12.901	1.3882	0.52813	16	0.07	017.42	0
017	Boron (mg / kg (ppm))	0918	13.120	0.12000	12.901	1.3882	0.52813	16	0.16	017.43	0
017	Boron (mg / kg (ppm))	0560	13.155	0.61000	12.901	1.3882	0.52813	16	0.18	017.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
017	Boron (mg / kg (ppm))	0229	13.915	0.37000	12.901	1.3882	0.52813	16	0.73	017.41	0
017	Boron (mg / kg (ppm))	2053	14.635	0.97000	12.901	1.3882	0.52813	16	1.25	017.42	0
017	Boron (mg / kg (ppm))	0510	16.000	0.00000	12.901	1.3882	0.52813	16	2.23	017.43	0
017	Boron (mg / kg (ppm))	0294	21.140	0.12000	12.901	1.3882	0.52813	16	5.93	017.42	0
019	Calcium (%)	0889	2.8858	0.15300	4.0261	0.21747	0.09081	147	-5.24	019.99	0
019	Calcium (%)	0596	3.1350	0.03000	4.0261	0.21747	0.09081	147	-4.10	019.31	0
019	Calcium (%)	0205	3.3050	0.03000	4.0261	0.21747	0.09081	147	-3.32	019.42	0
019	Calcium (%)	0590	3.3150	0.03000	4.0261	0.21747	0.09081	147	-3.27	019.99	0
019	Calcium (%)	2146	3.4600	0.09600	4.0261	0.21747	0.09081	147	-2.60	019.31	0
019	Calcium (%)	0042	3.4800	0.24000	4.0261	0.21747	0.09081	147	-2.51	019.42	0
019	Calcium (%)	0722	3.4811	0.08670	4.0261	0.21747	0.09081	147	-2.51	019.31	0
019	Calcium (%)	0038	3.5150	0.13000	4.0261	0.21747	0.09081	147	-2.35	019.43	0
019	Calcium (%)	0884	3.6100	0.06000	4.0261	0.21747	0.09081	147	-1.91	019.31	0
019	Calcium (%)	0676	3.6550	0.01000	4.0261	0.21747	0.09081	147	-1.71	019.99	0
019	Calcium (%)	0144	3.6800	0.02000	4.0261	0.21747	0.09081	147	-1.59	019.41	0
019	Calcium (%)	2192	3.6800	0.04000	4.0261	0.21747	0.09081	147	-1.59	019.43	0
019	Calcium (%)	0861	3.7150	0.43000	4.0261	0.21747	0.09081	147	-1.43	019.43	0
019	Calcium (%)	0263	3.7304	0.03160	4.0261	0.21747	0.09081	147	-1.36	019.42	0
019	Calcium (%)	2128	3.7320	0.32800	4.0261	0.21747	0.09081	147	-1.35	019.02	0
019	Calcium (%)	0298	3.7600	0.02000	4.0261	0.21747	0.09081	147	-1.22	019.41	0
019	Calcium (%)	0511	3.7650	0.05000	4.0261	0.21747	0.09081	147	-1.20	019.41	0
019	Calcium (%)	2129	3.7750	0.17000	4.0261	0.21747	0.09081	147	-1.15	019.42	0
019	Calcium (%)	0018	3.7750	0.03000	4.0261	0.21747	0.09081	147	-1.15	019.43	0
019	Calcium (%)	0622	3.7760	0.09250	4.0261	0.21747	0.09081	147	-1.15	019.00	0
019	Calcium (%)	0345	3.7850	0.05000	4.0261	0.21747	0.09081	147	-1.11	019.43	0
019	Calcium (%)	0045	3.7900	0.04000	4.0261	0.21747	0.09081	147	-1.09	019.42	0
019	Calcium (%)	0202	3.8040	0.40200	4.0261	0.21747	0.09081	147	-1.02	019.43	0
019	Calcium (%)	0610	3.8150	0.09000	4.0261	0.21747	0.09081	147	-0.97	019.43	0
019	Calcium (%)	0265	3.8250	0.27000	4.0261	0.21747	0.09081	147	-0.92	019.42	0
019	Calcium (%)	0563	3.8311	0.02770	4.0261	0.21747	0.09081	147	-0.90	019.31	0
019	Calcium (%)	0036	3.8313	0.11090	4.0261	0.21747	0.09081	147	-0.90	019.43	0
019	Calcium (%)	0171	3.8400	0.02000	4.0261	0.21747	0.09081	147	-0.86	019.41	0
019	Calcium (%)	0629	3.8450	0.01000	4.0261	0.21747	0.09081	147	-0.83	019.08	0
019	Calcium (%)	0848	3.8450	0.01000	4.0261	0.21747	0.09081	147	-0.83	019.41	0
019	Calcium (%)	0038	3.8540	0.19800	4.0261	0.21747	0.09081	147	-0.79	019.31	0
019	Calcium (%)	0504	3.8595	0.22700	4.0261	0.21747	0.09081	147	-0.77	019.42	0
019	Calcium (%)	0590	3.8600	0.12000	4.0261	0.21747	0.09081	147	-0.76	019.08	0
019	Calcium (%)	0278	3.8600	0.02000	4.0261	0.21747	0.09081	147	-0.76	019.42	0
019	Calcium (%)	2181	3.8650	0.05000	4.0261	0.21747	0.09081	147	-0.74	019.00	0
019	Calcium (%)	0939	3.8850	0.07000	4.0261	0.21747	0.09081	147	-0.65	019.31	0
019	Calcium (%)	0164	3.8850	0.25000	4.0261	0.21747	0.09081	147	-0.65	019.41	0
019	Calcium (%)	0043	3.8850	0.13000	4.0261	0.21747	0.09081	147	-0.65	019.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	0098	3.8900	0.00000	4.0261	0.21747	0.09081	147	-0.63	019.41	0
019	Calcium (%)	0874	3.9000	0.10000	4.0261	0.21747	0.09081	147	-0.58	019.31	0
019	Calcium (%)	0849	3.9010	0.02200	4.0261	0.21747	0.09081	147	-0.58	019.08	0
019	Calcium (%)	0940	3.9100	0.06000	4.0261	0.21747	0.09081	147	-0.53	019.08	0
019	Calcium (%)	0010	3.9100	0.16000	4.0261	0.21747	0.09081	147	-0.53	019.43	0
019	Calcium (%)	0208	3.9140	0.34000	4.0261	0.21747	0.09081	147	-0.52	019.41	0
019	Calcium (%)	0651	3.9200	0.18000	4.0261	0.21747	0.09081	147	-0.49	019.00	0
019	Calcium (%)	2144	3.9200	0.02000	4.0261	0.21747	0.09081	147	-0.49	019.99	0
019	Calcium (%)	0536	3.9250	0.01000	4.0261	0.21747	0.09081	147	-0.46	019.31	0
019	Calcium (%)	0885	3.9375	0.03630	4.0261	0.21747	0.09081	147	-0.41	019.08	0
019	Calcium (%)	0019	3.9400	0.08000	4.0261	0.21747	0.09081	147	-0.40	019.41	0
019	Calcium (%)	0968	3.9425	0.00300	4.0261	0.21747	0.09081	147	-0.38	019.43	0
019	Calcium (%)	0852	3.9500	0.10000	4.0261	0.21747	0.09081	147	-0.35	019.08	0
019	Calcium (%)	2022	3.9500	0.02000	4.0261	0.21747	0.09081	147	-0.35	019.31	0
019	Calcium (%)	0208	3.9520	0.12000	4.0261	0.21747	0.09081	147	-0.34	019.31	0
019	Calcium (%)	0910	3.9550	0.05000	4.0261	0.21747	0.09081	147	-0.33	019.41	0
019	Calcium (%)	0035	3.9550	0.11000	4.0261	0.21747	0.09081	147	-0.33	019.42	0
019	Calcium (%)	0870	3.9580	0.01600	4.0261	0.21747	0.09081	147	-0.31	019.42	0
019	Calcium (%)	0897	3.9600	0.02000	4.0261	0.21747	0.09081	147	-0.30	019.00	0
019	Calcium (%)	0675	3.9600	0.04000	4.0261	0.21747	0.09081	147	-0.30	019.43	0
019	Calcium (%)	0148	3.9700	0.04000	4.0261	0.21747	0.09081	147	-0.26	019.41	0
019	Calcium (%)	0175	3.9750	0.09000	4.0261	0.21747	0.09081	147	-0.23	019.31	0
019	Calcium (%)	0089	3.9750	0.01000	4.0261	0.21747	0.09081	147	-0.23	019.43	0
019	Calcium (%)	0242	3.9900	0.00000	4.0261	0.21747	0.09081	147	-0.17	019.99	0
019	Calcium (%)	0139	3.9930	0.04000	4.0261	0.21747	0.09081	147	-0.15	019.41	0
019	Calcium (%)	0687	3.9950	0.21000	4.0261	0.21747	0.09081	147	-0.14	019.31	0
019	Calcium (%)	0878	4.0000	0.20000	4.0261	0.21747	0.09081	147	-0.12	019.41	0
019	Calcium (%)	0035	4.0050	0.05000	4.0261	0.21747	0.09081	147	-0.10	019.43	0
019	Calcium (%)	2109	4.0155	0.05500	4.0261	0.21747	0.09081	147	-0.05	019.00	0
019	Calcium (%)	0034	4.0200	0.30200	4.0261	0.21747	0.09081	147	-0.03	019.43	0
019	Calcium (%)	2145	4.0300	0.06000	4.0261	0.21747	0.09081	147	0.02	019.41	0
019	Calcium (%)	0668	4.0300	0.02000	4.0261	0.21747	0.09081	147	0.02	019.43	0
019	Calcium (%)	2193	4.0300	0.00000	4.0261	0.21747	0.09081	147	0.02	019.08	0
019	Calcium (%)	2006	4.0320	0.01000	4.0261	0.21747	0.09081	147	0.03	019.00	0
019	Calcium (%)	2009	4.0354	0.00540	4.0261	0.21747	0.09081	147	0.04	019.08	0
019	Calcium (%)	0017	4.0430	0.08400	4.0261	0.21747	0.09081	147	0.08	019.43	0
019	Calcium (%)	2188	4.0450	0.05000	4.0261	0.21747	0.09081	147	0.09	019.08	0
019	Calcium (%)	0354	4.0450	0.01000	4.0261	0.21747	0.09081	147	0.09	019.31	0
019	Calcium (%)	0229	4.0450	0.15000	4.0261	0.21747	0.09081	147	0.09	019.41	0
019	Calcium (%)	0186	4.0450	0.07000	4.0261	0.21747	0.09081	147	0.09	019.42	0
019	Calcium (%)	0142	4.0497	0.00650	4.0261	0.21747	0.09081	147	0.11	019.31	0
019	Calcium (%)	0074	4.0500	0.10000	4.0261	0.21747	0.09081	147	0.11	019.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	0692	4.0500	0.02000	4.0261	0.21747	0.09081	147	0.11	019.42	0
019	Calcium (%)	0047	4.0500	0.06000	4.0261	0.21747	0.09081	147	0.11	019.52	0
019	Calcium (%)	0683	4.0500	0.12000	4.0261	0.21747	0.09081	147	0.11	019.00	0
019	Calcium (%)	2195	4.0600	0.04000	4.0261	0.21747	0.09081	147	0.16	019.08	0
019	Calcium (%)	0037	4.0600	0.06000	4.0261	0.21747	0.09081	147	0.16	019.42	0
019	Calcium (%)	0878	4.0650	0.03000	4.0261	0.21747	0.09081	147	0.18	019.00	0
019	Calcium (%)	2006	4.0695	0.03700	4.0261	0.21747	0.09081	147	0.20	019.09	0
019	Calcium (%)	2196	4.0700	0.00000	4.0261	0.21747	0.09081	147	0.20	019.31	0
019	Calcium (%)	0510	4.0700	0.16000	4.0261	0.21747	0.09081	147	0.20	019.43	0
019	Calcium (%)	0169	4.0800	0.02000	4.0261	0.21747	0.09081	147	0.25	019.32	0
019	Calcium (%)	0006	4.0970	0.18600	4.0261	0.21747	0.09081	147	0.33	019.43	0
019	Calcium (%)	2196	4.1000	0.00000	4.0261	0.21747	0.09081	147	0.34	019.08	0
019	Calcium (%)	0187	4.1000	0.02000	4.0261	0.21747	0.09081	147	0.34	019.42	0
019	Calcium (%)	0964	4.1019	0.08490	4.0261	0.21747	0.09081	147	0.35	019.43	0
019	Calcium (%)	0895	4.1050	0.01000	4.0261	0.21747	0.09081	147	0.36	019.00	0
019	Calcium (%)	2188	4.1100	0.02000	4.0261	0.21747	0.09081	147	0.39	019.31	0
019	Calcium (%)	0689	4.1100	0.00000	4.0261	0.21747	0.09081	147	0.39	019.08	0
019	Calcium (%)	0036	4.1129	0.06340	4.0261	0.21747	0.09081	147	0.40	019.03	0
019	Calcium (%)	0405	4.1150	0.01000	4.0261	0.21747	0.09081	147	0.41	019.41	0
019	Calcium (%)	0098	4.1150	0.11000	4.0261	0.21747	0.09081	147	0.41	019.43	0
019	Calcium (%)	2190	4.1158	0.14520	4.0261	0.21747	0.09081	147	0.41	019.08	0
019	Calcium (%)	0294	4.1250	0.07000	4.0261	0.21747	0.09081	147	0.45	019.42	0
019	Calcium (%)	0881	4.1275	0.07100	4.0261	0.21747	0.09081	147	0.47	019.00	0
019	Calcium (%)	0598	4.1330	0.05400	4.0261	0.21747	0.09081	147	0.49	019.41	0
019	Calcium (%)	0049	4.1400	0.20000	4.0261	0.21747	0.09081	147	0.52	019.41	0
019	Calcium (%)	0653	4.1425	0.08700	4.0261	0.21747	0.09081	147	0.54	019.41	0
019	Calcium (%)	0650	4.1450	0.37000	4.0261	0.21747	0.09081	147	0.55	019.31	0
019	Calcium (%)	0358	4.1500	0.10000	4.0261	0.21747	0.09081	147	0.57	019.41	0
019	Calcium (%)	0413	4.1500	0.30000	4.0261	0.21747	0.09081	147	0.57	019.41	0
019	Calcium (%)	0553	4.1500	0.12000	4.0261	0.21747	0.09081	147	0.57	019.53	0
019	Calcium (%)	0297	4.1550	0.03000	4.0261	0.21747	0.09081	147	0.59	019.43	0
019	Calcium (%)	0003	4.1550	0.15000	4.0261	0.21747	0.09081	147	0.59	019.41	0
019	Calcium (%)	0357	4.1602	0.04880	4.0261	0.21747	0.09081	147	0.62	019.42	0
019	Calcium (%)	0226	4.1700	0.12000	4.0261	0.21747	0.09081	147	0.66	019.41	0
019	Calcium (%)	0425	4.1700	0.02000	4.0261	0.21747	0.09081	147	0.66	019.43	0
019	Calcium (%)	2089	4.1700	0.18000	4.0261	0.21747	0.09081	147	0.66	019.43	0
019	Calcium (%)	0629	4.1750	0.01000	4.0261	0.21747	0.09081	147	0.68	019.43	0
019	Calcium (%)	0504	4.1965	0.12900	4.0261	0.21747	0.09081	147	0.78	019.32	0
019	Calcium (%)	0520	4.2000	0.18000	4.0261	0.21747	0.09081	147	0.80	019.41	0
019	Calcium (%)	0720	4.2050	0.05000	4.0261	0.21747	0.09081	147	0.82	019.41	0
019	Calcium (%)	0693	4.2150	0.01000	4.0261	0.21747	0.09081	147	0.87	019.42	0
019	Calcium (%)	0948	4.2155	0.02100	4.0261	0.21747	0.09081	147	0.87	019.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	2161	4.2250	0.07000	4.0261	0.21747	0.09081	147	0.91	019.99	0
019	Calcium (%)	0027	4.2335	0.05100	4.0261	0.21747	0.09081	147	0.95	019.43	0
019	Calcium (%)	0154	4.2356	0.22460	4.0261	0.21747	0.09081	147	0.96	019.52	0
019	Calcium (%)	0001	4.2405	0.11900	4.0261	0.21747	0.09081	147	0.99	019.31	0
019	Calcium (%)	2053	4.2700	0.20000	4.0261	0.21747	0.09081	147	1.12	019.42	0
019	Calcium (%)	0108	4.2750	0.11000	4.0261	0.21747	0.09081	147	1.14	019.31	0
019	Calcium (%)	0682	4.2800	0.00000	4.0261	0.21747	0.09081	147	1.17	019.43	0
019	Calcium (%)	2199	4.2900	0.02000	4.0261	0.21747	0.09081	147	1.21	019.08	0
019	Calcium (%)	0300	4.2900	0.08000	4.0261	0.21747	0.09081	147	1.21	019.33	0
019	Calcium (%)	0609	4.2950	0.09000	4.0261	0.21747	0.09081	147	1.24	019.32	0
019	Calcium (%)	0407	4.3533	0.16080	4.0261	0.21747	0.09081	147	1.50	019.41	0
019	Calcium (%)	2126	4.3600	0.00000	4.0261	0.21747	0.09081	147	1.54	019.99	0
019	Calcium (%)	0868	4.3750	0.01000	4.0261	0.21747	0.09081	147	1.60	019.31	0
019	Calcium (%)	0190	4.3800	0.06000	4.0261	0.21747	0.09081	147	1.63	019.42	0
019	Calcium (%)	0100	4.3850	0.05000	4.0261	0.21747	0.09081	147	1.65	019.99	0
019	Calcium (%)	0033	4.3900	0.08000	4.0261	0.21747	0.09081	147	1.67	019.43	0
019	Calcium (%)	0505	4.3900	0.12000	4.0261	0.21747	0.09081	147	1.67	019.41	0
019	Calcium (%)	0043	4.4000	0.20000	4.0261	0.21747	0.09081	147	1.72	019.00	0
019	Calcium (%)	0186	4.4000	0.00000	4.0261	0.21747	0.09081	147	1.72	019.52	0
019	Calcium (%)	0032	4.4250	0.31000	4.0261	0.21747	0.09081	147	1.83	019.43	0
019	Calcium (%)	0726	4.4930	0.00400	4.0261	0.21747	0.09081	147	2.15	019.42	0
019	Calcium (%)	0083	4.5800	0.16000	4.0261	0.21747	0.09081	147	2.55	019.43	0
019	Calcium (%)	0353	4.6200	0.20000	4.0261	0.21747	0.09081	147	2.73	019.43	0
019	Calcium (%)	0536	4.8600	0.10000	4.0261	0.21747	0.09081	147	3.83	019.02	0
019	Calcium (%)	0918	5.0515	0.18900	4.0261	0.21747	0.09081	147	4.72	019.43	0
019	Calcium (%)	2200	2.6500	0.50000	4.0261	0.21747	0.09081	147	-6.33	019.08	1
019	Calcium (%)	0560	4.0661	0.56370	4.0261	0.21747	0.09081	147	0.18	019.42	1
019	Calcium (%)	0572	4.1100	0.64000	4.0261	0.21747	0.09081	147	0.39	019.53	1
021	Cobalt (mg / kg (ppm))	0693	0.67250	0.01100	1.6222	0.37304	0.08482	18	-2.55	021.42	0
021	Cobalt (mg / kg (ppm))	0910	0.97500	0.03000	1.6222	0.37304	0.08482	18	-1.73	021.52	0
021	Cobalt (mg / kg (ppm))	0345	1.2350	0.03000	1.6222	0.37304	0.08482	18	-1.04	021.43	0
021	Cobalt (mg / kg (ppm))	0918	1.4161	0.05120	1.6222	0.37304	0.08482	18	-0.55	021.53	0
021	Cobalt (mg / kg (ppm))	0186	1.4300	0.10000	1.6222	0.37304	0.08482	18	-0.52	021.52	0
021	Cobalt (mg / kg (ppm))	0689	1.4500	0.10000	1.6222	0.37304	0.08482	18	-0.46	021.31	0
021	Cobalt (mg / kg (ppm))	0171	1.5080	0.02200	1.6222	0.37304	0.08482	18	-0.31	021.41	0
021	Cobalt (mg / kg (ppm))	0045	1.5300	0.08000	1.6222	0.37304	0.08482	18	-0.25	021.42	0
021	Cobalt (mg / kg (ppm))	0560	1.5700	0.00000	1.6222	0.37304	0.08482	18	-0.14	021.42	0
021	Cobalt (mg / kg (ppm))	0043	1.5850	0.03000	1.6222	0.37304	0.08482	18	-0.10	021.43	0
021	Cobalt (mg / kg (ppm))	0098	1.6760	0.05600	1.6222	0.37304	0.08482	18	0.14	021.53	0
021	Cobalt (mg / kg (ppm))	0169	1.7400	0.00000	1.6222	0.37304	0.08482	18	0.32	021.43	0
021	Cobalt (mg / kg (ppm))	0964	1.7972	0.19060	1.6222	0.37304	0.08482	18	0.47	021.43	0
021	Cobalt (mg / kg (ppm))	0553	1.8500	0.22000	1.6222	0.37304	0.08482	18	0.61	021.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
021	Cobalt (mg / kg (ppm))	0510	1.9900	0.00000	1.6222	0.37304	0.08482	18	0.99	021.43	0
021	Cobalt (mg / kg (ppm))	0038	2.0250	0.15000	1.6222	0.37304	0.08482	18	1.08	021.43	0
021	Cobalt (mg / kg (ppm))	0939	2.0900	0.42000	1.6222	0.37304	0.08482	18	1.25	021.31	0
021	Cobalt (mg / kg (ppm))	0032	2.3770	0.03600	1.6222	0.37304	0.08482	18	2.02	021.53	0
021	Cobalt (mg / kg (ppm))	0148	2.3750	0.53000	1.6222	0.37304	0.08482	18	2.02	021.41	1
021	Cobalt (mg / kg (ppm))	2161	< 3.5		1.6222	0.37304	0.08482	18		021.99	5
022	Copper (mg / kg (ppm))	2161	53.300	1.2000	190.74	17.571	6.2969	99	-7.82	022.99	0
022	Copper (mg / kg (ppm))	0405	71.350	1.1000	190.74	17.571	6.2969	99	-6.79	022.41	0
022	Copper (mg / kg (ppm))	0037	110.75	20.300	190.74	17.571	6.2969	99	-4.55	022.42	0
022	Copper (mg / kg (ppm))	0596	148.10	3.2000	190.74	17.571	6.2969	99	-2.43	022.31	0
022	Copper (mg / kg (ppm))	0049	152.16	7.2700	190.74	17.571	6.2969	99	-2.20	022.41	0
022	Copper (mg / kg (ppm))	0720	154.33	4.3000	190.74	17.571	6.2969	99	-2.07	022.41	0
022	Copper (mg / kg (ppm))	0047	155.10	8.0000	190.74	17.571	6.2969	99	-2.03	022.52	0
022	Copper (mg / kg (ppm))	0353	162.50	7.0000	190.74	17.571	6.2969	99	-1.61	022.43	0
022	Copper (mg / kg (ppm))	0098	165.55	12.500	190.74	17.571	6.2969	99	-1.43	022.41	0
022	Copper (mg / kg (ppm))	0358	166.99	5.8200	190.74	17.571	6.2969	99	-1.35	022.41	0
022	Copper (mg / kg (ppm))	0520	167.00	0.00000	190.74	17.571	6.2969	99	-1.35	022.41	0
022	Copper (mg / kg (ppm))	0505	168.00	1.6000	190.74	17.571	6.2969	99	-1.29	022.41	0
022	Copper (mg / kg (ppm))	0001	168.80	17.600	190.74	17.571	6.2969	99	-1.25	022.31	0
022	Copper (mg / kg (ppm))	0345	172.00	2.0000	190.74	17.571	6.2969	99	-1.07	022.43	0
022	Copper (mg / kg (ppm))	0910	172.50	23.000	190.74	17.571	6.2969	99	-1.04	022.41	0
022	Copper (mg / kg (ppm))	0590	172.50	1.0000	190.74	17.571	6.2969	99	-1.04	022.99	0
022	Copper (mg / kg (ppm))	0653	173.05	2.5000	190.74	17.571	6.2969	99	-1.01	022.41	0
022	Copper (mg / kg (ppm))	0045	174.50	5.0000	190.74	17.571	6.2969	99	-0.92	022.42	0
022	Copper (mg / kg (ppm))	0563	174.52	5.7896	190.74	17.571	6.2969	99	-0.92	022.31	0
022	Copper (mg / kg (ppm))	0536	174.80	1.9000	190.74	17.571	6.2969	99	-0.91	022.31	0
022	Copper (mg / kg (ppm))	0609	175.00	0.00000	190.74	17.571	6.2969	99	-0.90	022.32	0
022	Copper (mg / kg (ppm))	0202	175.24	2.4600	190.74	17.571	6.2969	99	-0.88	022.43	0
022	Copper (mg / kg (ppm))	0148	175.25	1.7000	190.74	17.571	6.2969	99	-0.88	022.41	0
022	Copper (mg / kg (ppm))	2089	176.04	9.9200	190.74	17.571	6.2969	99	-0.84	022.43	0
022	Copper (mg / kg (ppm))	0878	178.00	12.000	190.74	17.571	6.2969	99	-0.73	022.41	0
022	Copper (mg / kg (ppm))	0407	179.17	2.4459	190.74	17.571	6.2969	99	-0.66	022.41	0
022	Copper (mg / kg (ppm))	0689	179.75	7.5000	190.74	17.571	6.2969	99	-0.63	022.31	0
022	Copper (mg / kg (ppm))	0278	180.00	8.0000	190.74	17.571	6.2969	99	-0.61	022.42	0
022	Copper (mg / kg (ppm))	0186	180.50	3.0000	190.74	17.571	6.2969	99	-0.58	022.52	0
022	Copper (mg / kg (ppm))	0019	181.15	0.50000	190.74	17.571	6.2969	99	-0.55	022.41	0
022	Copper (mg / kg (ppm))	0598	183.25	2.7000	190.74	17.571	6.2969	99	-0.43	022.41	0
022	Copper (mg / kg (ppm))	0948	183.89	0.51900	190.74	17.571	6.2969	99	-0.39	022.33	0
022	Copper (mg / kg (ppm))	0354	184.67	1.3000	190.74	17.571	6.2969	99	-0.35	022.31	0
022	Copper (mg / kg (ppm))	0038	185.25	12.700	190.74	17.571	6.2969	99	-0.31	022.32	0
022	Copper (mg / kg (ppm))	0175	186.00	8.0000	190.74	17.571	6.2969	99	-0.27	022.31	0
022	Copper (mg / kg (ppm))	2195	186.50	1.0000	190.74	17.571	6.2969	99	-0.24	022.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
022	Copper (mg / kg (ppm))	0164	186.50	15.000	190.74	17.571	6.2969	99	-0.24	022.41	0
022	Copper (mg / kg (ppm))	0229	187.15	0.50000	190.74	17.571	6.2969	99	-0.20	022.41	0
022	Copper (mg / kg (ppm))	0265	187.50	21.000	190.74	17.571	6.2969	99	-0.18	022.42	0
022	Copper (mg / kg (ppm))	0968	187.50	3.0000	190.74	17.571	6.2969	99	-0.18	022.43	0
022	Copper (mg / kg (ppm))	0208	187.60	6.0000	190.74	17.571	6.2969	99	-0.18	022.41	0
022	Copper (mg / kg (ppm))	0043	188.50	13.000	190.74	17.571	6.2969	99	-0.13	022.43	0
022	Copper (mg / kg (ppm))	0171	188.75	3.5000	190.74	17.571	6.2969	99	-0.11	022.41	0
022	Copper (mg / kg (ppm))	0074	189.50	3.0000	190.74	17.571	6.2969	99	-0.07	022.41	0
022	Copper (mg / kg (ppm))	2188	190.06	1.3735	190.74	17.571	6.2969	99	-0.04	022.31	0
022	Copper (mg / kg (ppm))	2129	190.70	0.60000	190.74	17.571	6.2969	99	0.00	022.99	0
022	Copper (mg / kg (ppm))	2196	190.80	0.00000	190.74	17.571	6.2969	99	0.00	022.31	0
022	Copper (mg / kg (ppm))	2022	191.00	6.0000	190.74	17.571	6.2969	99	0.01	022.31	0
022	Copper (mg / kg (ppm))	0868	191.35	7.0100	190.74	17.571	6.2969	99	0.03	022.31	0
022	Copper (mg / kg (ppm))	0726	191.35	2.3000	190.74	17.571	6.2969	99	0.03	022.42	0
022	Copper (mg / kg (ppm))	0190	191.48	0.36000	190.74	17.571	6.2969	99	0.04	022.42	0
022	Copper (mg / kg (ppm))	2144	191.86	8.6900	190.74	17.571	6.2969	99	0.06	022.99	0
022	Copper (mg / kg (ppm))	0610	192.00	2.0000	190.74	17.571	6.2969	99	0.07	022.43	0
022	Copper (mg / kg (ppm))	2145	192.70	0.60000	190.74	17.571	6.2969	99	0.11	022.41	0
022	Copper (mg / kg (ppm))	0668	193.00	4.0000	190.74	17.571	6.2969	99	0.13	022.43	0
022	Copper (mg / kg (ppm))	0511	193.50	7.0000	190.74	17.571	6.2969	99	0.16	022.41	0
022	Copper (mg / kg (ppm))	0098	194.90	1.8000	190.74	17.571	6.2969	99	0.24	022.43	0
022	Copper (mg / kg (ppm))	0205	195.50	1.0000	190.74	17.571	6.2969	99	0.27	022.42	0
022	Copper (mg / kg (ppm))	0208	196.00	2.0000	190.74	17.571	6.2969	99	0.30	022.31	0
022	Copper (mg / kg (ppm))	0939	196.77	3.8000	190.74	17.571	6.2969	99	0.34	022.31	0
022	Copper (mg / kg (ppm))	0033	197.50	21.000	190.74	17.571	6.2969	99	0.38	022.43	0
022	Copper (mg / kg (ppm))	0964	198.61	10.290	190.74	17.571	6.2969	99	0.45	022.43	0
022	Copper (mg / kg (ppm))	0187	198.73	0.94000	190.74	17.571	6.2969	99	0.45	022.42	0
022	Copper (mg / kg (ppm))	0017	198.87	4.6460	190.74	17.571	6.2969	99	0.46	022.43	0
022	Copper (mg / kg (ppm))	0032	199.45	7.3000	190.74	17.571	6.2969	99	0.50	022.43	0
022	Copper (mg / kg (ppm))	0553	199.50	1.0000	190.74	17.571	6.2969	99	0.50	022.53	0
022	Copper (mg / kg (ppm))	0861	200.00	6.0000	190.74	17.571	6.2969	99	0.53	022.43	0
022	Copper (mg / kg (ppm))	0884	200.10	3.4000	190.74	17.571	6.2969	99	0.53	022.31	0
022	Copper (mg / kg (ppm))	0035	200.50	1.0000	190.74	17.571	6.2969	99	0.56	022.42	0
022	Copper (mg / kg (ppm))	0242	200.50	9.0000	190.74	17.571	6.2969	99	0.56	022.99	0
022	Copper (mg / kg (ppm))	0682	200.60	0.00000	190.74	17.571	6.2969	99	0.56	022.43	0
022	Copper (mg / kg (ppm))	2128	200.72	26.532	190.74	17.571	6.2969	99	0.57	022.32	0
022	Copper (mg / kg (ppm))	0035	201.00	8.0000	190.74	17.571	6.2969	99	0.58	022.43	0
022	Copper (mg / kg (ppm))	0100	201.00	6.0000	190.74	17.571	6.2969	99	0.58	022.99	0
022	Copper (mg / kg (ppm))	0675	201.95	3.7600	190.74	17.571	6.2969	99	0.64	022.43	0
022	Copper (mg / kg (ppm))	0294	202.65	8.3000	190.74	17.571	6.2969	99	0.68	022.42	0
022	Copper (mg / kg (ppm))	0685	202.74	1.6200	190.74	17.571	6.2969	99	0.68	022.42	0
022	Copper (mg / kg (ppm))	0692	203.00	8.0000	190.74	17.571	6.2969	99	0.70	022.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
022	Copper (mg / kg (ppm))	0357	203.87	0.70900	190.74	17.571	6.2969	99	0.75	022.42	0
022	Copper (mg / kg (ppm))	0186	204.00	4.0000	190.74	17.571	6.2969	99	0.75	022.42	0
022	Copper (mg / kg (ppm))	0010	204.50	13.800	190.74	17.571	6.2969	99	0.78	022.33	0
022	Copper (mg / kg (ppm))	0504	205.00	3.6000	190.74	17.571	6.2969	99	0.81	022.32	0
022	Copper (mg / kg (ppm))	0297	206.50	7.0000	190.74	17.571	6.2969	99	0.90	022.43	0
022	Copper (mg / kg (ppm))	0560	206.55	8.7000	190.74	17.571	6.2969	99	0.90	022.42	0
022	Copper (mg / kg (ppm))	0870	207.15	13.700	190.74	17.571	6.2969	99	0.93	022.42	0
022	Copper (mg / kg (ppm))	0042	208.50	11.000	190.74	17.571	6.2969	99	1.01	022.42	0
022	Copper (mg / kg (ppm))	0918	209.42	0.80000	190.74	17.571	6.2969	99	1.06	022.43	0
022	Copper (mg / kg (ppm))	0425	209.85	0.50000	190.74	17.571	6.2969	99	1.09	022.43	0
022	Copper (mg / kg (ppm))	0510	210.00	2.0000	190.74	17.571	6.2969	99	1.10	022.43	0
022	Copper (mg / kg (ppm))	0083	212.00	4.0000	190.74	17.571	6.2969	99	1.21	022.43	0
022	Copper (mg / kg (ppm))	0693	214.00	8.0000	190.74	17.571	6.2969	99	1.32	022.42	0
022	Copper (mg / kg (ppm))	0413	215.50	11.000	190.74	17.571	6.2969	99	1.41	022.42	0
022	Copper (mg / kg (ppm))	2053	215.50	3.0000	190.74	17.571	6.2969	99	1.41	022.42	0
022	Copper (mg / kg (ppm))	0027	217.42	1.3070	190.74	17.571	6.2969	99	1.52	022.43	0
022	Copper (mg / kg (ppm))	0226	217.50	41.000	190.74	17.571	6.2969	99	1.52	022.41	0
022	Copper (mg / kg (ppm))	0572	218.00	34.000	190.74	17.571	6.2969	99	1.55	022.53	0
022	Copper (mg / kg (ppm))	0038	218.39	3.8800	190.74	17.571	6.2969	99	1.57	022.43	0
022	Copper (mg / kg (ppm))	0169	224.00	2.0000	190.74	17.571	6.2969	99	1.89	022.43	0
022	Copper (mg / kg (ppm))	0263	236.76	4.2500	190.74	17.571	6.2969	99	2.62	022.42	0
022	Copper (mg / kg (ppm))	0003	122.50	91.000	190.74	17.571	6.2969	99	-3.88	022.41	1
023	Fluorine (mg / kg (ppm))	0910	58.700	0.40000				1		023.01	0
024	Iodine (mg / kg (ppm))	0171	2.7000	0.08000				2		024.03	0
024	Iodine (mg / kg (ppm))	0160	3.5550	0.59000				2		024.99	0
025	Iron (mg / kg (ppm))	0148	120.75	1.8100	1,181.5	104.77	35.504	87	-10.12	025.41	0
025	Iron (mg / kg (ppm))	0278	207.50	11.000	1,181.5	104.77	35.504	87	-9.30	025.42	0
025	Iron (mg / kg (ppm))	0609	330.00	0.00000	1,181.5	104.77	35.504	87	-8.13	025.32	0
025	Iron (mg / kg (ppm))	0405	590.70	2.6000	1,181.5	104.77	35.504	87	-5.64	025.41	0
025	Iron (mg / kg (ppm))	0726	661.65	5.7000	1,181.5	104.77	35.504	87	-4.96	025.42	0
025	Iron (mg / kg (ppm))	0047	794.65	3.5000	1,181.5	104.77	35.504	87	-3.69	025.52	0
025	Iron (mg / kg (ppm))	2161	834.00	60.000	1,181.5	104.77	35.504	87	-3.32	025.99	0
025	Iron (mg / kg (ppm))	0345	841.50	83.000	1,181.5	104.77	35.504	87	-3.25	025.43	0
025	Iron (mg / kg (ppm))	0043	963.50	43.000	1,181.5	104.77	35.504	87	-2.08	025.43	0
025	Iron (mg / kg (ppm))	0629	1,014.0	8.0000	1,181.5	104.77	35.504	87	-1.60	025.31	0
025	Iron (mg / kg (ppm))	2129	1,015.1	75.900	1,181.5	104.77	35.504	87	-1.59	025.99	0
025	Iron (mg / kg (ppm))	0169	1,018.0	124.00	1,181.5	104.77	35.504	87	-1.56	025.43	0
025	Iron (mg / kg (ppm))	0598	1,029.5	47.000	1,181.5	104.77	35.504	87	-1.45	025.41	0
025	Iron (mg / kg (ppm))	0560	1,066.7	135.80	1,181.5	104.77	35.504	87	-1.10	025.42	0
025	Iron (mg / kg (ppm))	0353	1,082.5	15.000	1,181.5	104.77	35.504	87	-0.95	025.43	0
025	Iron (mg / kg (ppm))	0870	1,086.5	17.000	1,181.5	104.77	35.504	87	-0.91	025.42	0
025	Iron (mg / kg (ppm))	0720	1,097.4	2.8000	1,181.5	104.77	35.504	87	-0.80	025.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
025	Iron (mg / kg (ppm))	0692	1,108.0	208.00	1,181.5	104.77	35.504	87	-0.70	025.42	0
025	Iron (mg / kg (ppm))	2128	1,108.6	113.42	1,181.5	104.77	35.504	87	-0.70	025.32	0
025	Iron (mg / kg (ppm))	0878	1,120.0	44.000	1,181.5	104.77	35.504	87	-0.59	025.41	0
025	Iron (mg / kg (ppm))	0653	1,120.5	1.0000	1,181.5	104.77	35.504	87	-0.58	025.41	0
025	Iron (mg / kg (ppm))	0190	1,121.9	7.8000	1,181.5	104.77	35.504	87	-0.57	025.42	0
025	Iron (mg / kg (ppm))	2145	1,123.0	2.0000	1,181.5	104.77	35.504	87	-0.56	025.41	0
025	Iron (mg / kg (ppm))	0263	1,126.7	8.2900	1,181.5	104.77	35.504	87	-0.52	025.42	0
025	Iron (mg / kg (ppm))	0610	1,130.0	10.000	1,181.5	104.77	35.504	87	-0.49	025.43	0
025	Iron (mg / kg (ppm))	0407	1,130.4	20.676	1,181.5	104.77	35.504	87	-0.49	025.41	0
025	Iron (mg / kg (ppm))	0045	1,135.0	10.000	1,181.5	104.77	35.504	87	-0.44	025.42	0
025	Iron (mg / kg (ppm))	2144	1,136.9	33.650	1,181.5	104.77	35.504	87	-0.43	025.99	0
025	Iron (mg / kg (ppm))	0164	1,139.0	98.000	1,181.5	104.77	35.504	87	-0.41	025.41	0
025	Iron (mg / kg (ppm))	2089	1,144.0	90.150	1,181.5	104.77	35.504	87	-0.36	025.43	0
025	Iron (mg / kg (ppm))	0874	1,145.0	30.000	1,181.5	104.77	35.504	87	-0.35	025.31	0
025	Iron (mg / kg (ppm))	2022	1,152.0	86.000	1,181.5	104.77	35.504	87	-0.28	025.31	0
025	Iron (mg / kg (ppm))	0520	1,153.5	5.0000	1,181.5	104.77	35.504	87	-0.27	025.41	0
025	Iron (mg / kg (ppm))	0563	1,157.6	82.671	1,181.5	104.77	35.504	87	-0.23	025.31	0
025	Iron (mg / kg (ppm))	0689	1,159.3	68.800	1,181.5	104.77	35.504	87	-0.21	025.31	0
025	Iron (mg / kg (ppm))	0884	1,162.5	2.5000	1,181.5	104.77	35.504	87	-0.18	025.31	0
025	Iron (mg / kg (ppm))	0504	1,170.0	50.000	1,181.5	104.77	35.504	87	-0.11	025.32	0
025	Iron (mg / kg (ppm))	0098	1,174.5	19.000	1,181.5	104.77	35.504	87	-0.07	025.41	0
025	Iron (mg / kg (ppm))	0910	1,178.0	6.0000	1,181.5	104.77	35.504	87	-0.03	025.41	0
025	Iron (mg / kg (ppm))	0553	1,180.0	60.000	1,181.5	104.77	35.504	87	-0.01	025.53	0
025	Iron (mg / kg (ppm))	0034	1,189.0	18.000	1,181.5	104.77	35.504	87	0.07	025.43	0
025	Iron (mg / kg (ppm))	0968	1,189.5	5.0000	1,181.5	104.77	35.504	87	0.08	025.43	0
025	Iron (mg / kg (ppm))	0037	1,191.0	6.0000	1,181.5	104.77	35.504	87	0.09	025.42	0
025	Iron (mg / kg (ppm))	0033	1,195.0	10.000	1,181.5	104.77	35.504	87	0.13	025.43	0
025	Iron (mg / kg (ppm))	0171	1,199.5	9.0000	1,181.5	104.77	35.504	87	0.17	025.41	0
025	Iron (mg / kg (ppm))	0208	1,200.0	6.0000	1,181.5	104.77	35.504	87	0.18	025.41	0
025	Iron (mg / kg (ppm))	0242	1,203.5	33.000	1,181.5	104.77	35.504	87	0.21	025.99	0
025	Iron (mg / kg (ppm))	0354	1,203.9	1.2400	1,181.5	104.77	35.504	87	0.21	025.31	0
025	Iron (mg / kg (ppm))	0358	1,204.9	5.9700	1,181.5	104.77	35.504	87	0.22	025.41	0
025	Iron (mg / kg (ppm))	0939	1,205.0	8.0000	1,181.5	104.77	35.504	87	0.22	025.31	0
025	Iron (mg / kg (ppm))	0511	1,211.5	31.000	1,181.5	104.77	35.504	87	0.29	025.41	0
025	Iron (mg / kg (ppm))	0049	1,214.4	13.470	1,181.5	104.77	35.504	87	0.31	025.41	0
025	Iron (mg / kg (ppm))	0017	1,216.2	25.401	1,181.5	104.77	35.504	87	0.33	025.43	0
025	Iron (mg / kg (ppm))	0098	1,217.5	53.000	1,181.5	104.77	35.504	87	0.34	025.43	0
025	Iron (mg / kg (ppm))	2195	1,219.0	16.000	1,181.5	104.77	35.504	87	0.36	025.31	0
025	Iron (mg / kg (ppm))	0042	1,223.0	16.000	1,181.5	104.77	35.504	87	0.40	025.42	0
025	Iron (mg / kg (ppm))	0510	1,230.0	20.000	1,181.5	104.77	35.504	87	0.46	025.43	0
025	Iron (mg / kg (ppm))	0505	1,233.5	45.000	1,181.5	104.77	35.504	87	0.50	025.41	0
025	Iron (mg / kg (ppm))	2053	1,234.0	22.000	1,181.5	104.77	35.504	87	0.50	025.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
025	Iron (mg / kg (ppm))	0297	1,234.5	1.0000	1,181.5	104.77	35.504	87	0.51	025.43	0
025	Iron (mg / kg (ppm))	0964	1,237.9	44.050	1,181.5	104.77	35.504	87	0.54	025.43	0
025	Iron (mg / kg (ppm))	0425	1,241.8	12.600	1,181.5	104.77	35.504	87	0.58	025.43	0
025	Iron (mg / kg (ppm))	0294	1,242.4	47.460	1,181.5	104.77	35.504	87	0.58	025.42	0
025	Iron (mg / kg (ppm))	2196	1,242.6	0.00000	1,181.5	104.77	35.504	87	0.58	025.31	0
025	Iron (mg / kg (ppm))	0003	1,244.0	118.00	1,181.5	104.77	35.504	87	0.60	025.41	0
025	Iron (mg / kg (ppm))	0629	1,245.0	50.000	1,181.5	104.77	35.504	87	0.61	025.43	0
025	Iron (mg / kg (ppm))	0226	1,248.0	22.000	1,181.5	104.77	35.504	87	0.63	025.41	0
025	Iron (mg / kg (ppm))	2192	1,253.2	75.650	1,181.5	104.77	35.504	87	0.68	025.43	0
025	Iron (mg / kg (ppm))	0019	1,262.5	7.0000	1,181.5	104.77	35.504	87	0.77	025.41	0
025	Iron (mg / kg (ppm))	0682	1,266.2	0.00000	1,181.5	104.77	35.504	87	0.81	025.43	0
025	Iron (mg / kg (ppm))	0074	1,269.5	21.000	1,181.5	104.77	35.504	87	0.84	025.41	0
025	Iron (mg / kg (ppm))	0413	1,270.0	120.00	1,181.5	104.77	35.504	87	0.84	025.42	0
025	Iron (mg / kg (ppm))	0693	1,270.0	60.000	1,181.5	104.77	35.504	87	0.84	025.42	0
025	Iron (mg / kg (ppm))	0868	1,272.7	8.5000	1,181.5	104.77	35.504	87	0.87	025.31	0
025	Iron (mg / kg (ppm))	0229	1,294.5	5.0000	1,181.5	104.77	35.504	87	1.08	025.41	0
025	Iron (mg / kg (ppm))	0001	1,296.0	16.000	1,181.5	104.77	35.504	87	1.09	025.31	0
025	Iron (mg / kg (ppm))	0100	1,299.0	20.000	1,181.5	104.77	35.504	87	1.12	025.99	0
025	Iron (mg / kg (ppm))	0675	1,299.6	75.990	1,181.5	104.77	35.504	87	1.13	025.43	0
025	Iron (mg / kg (ppm))	0175	1,300.0	0.00000	1,181.5	104.77	35.504	87	1.13	025.31	0
025	Iron (mg / kg (ppm))	2188	1,304.9	11.349	1,181.5	104.77	35.504	87	1.18	025.31	0
025	Iron (mg / kg (ppm))	0918	1,305.5	63.830	1,181.5	104.77	35.504	87	1.18	025.43	0
025	Iron (mg / kg (ppm))	0038	1,311.5	39.000	1,181.5	104.77	35.504	87	1.24	025.31	0
025	Iron (mg / kg (ppm))	0032	1,314.0	46.000	1,181.5	104.77	35.504	87	1.26	025.43	0
025	Iron (mg / kg (ppm))	0208	1,370.0	60.000	1,181.5	104.77	35.504	87	1.80	025.31	0
025	Iron (mg / kg (ppm))	2178	1,410.5	19.000	1,181.5	104.77	35.504	87	2.19	025.33	0
025	Iron (mg / kg (ppm))	0083	1,465.5	29.000	1,181.5	104.77	35.504	87	2.71	025.43	0
025	Iron (mg / kg (ppm))	0536	1,569.7	79.300	1,181.5	104.77	35.504	87	3.71	025.31	0
025	Iron (mg / kg (ppm))	0265	797.50	475.00	1,181.5	104.77	35.504	87	-3.67	025.42	1
025	Iron (mg / kg (ppm))	0861	1,031.0	818.00	1,181.5	104.77	35.504	87	-1.44	025.43	1
027	Magnesium (%)	0692	0.18000	0.04000	0.21593	0.01136	0.00696	99	-3.16	027.42	0
027	Magnesium (%)	0407	0.18310	0.00700	0.21593	0.01136	0.00696	99	-2.89	027.41	0
027	Magnesium (%)	0668	0.18650	0.00700	0.21593	0.01136	0.00696	99	-2.59	027.43	0
027	Magnesium (%)	2129	0.18700	0.00200	0.21593	0.01136	0.00696	99	-2.55	027.42	0
027	Magnesium (%)	0345	0.19150	0.00300	0.21593	0.01136	0.00696	99	-2.15	027.43	0
027	Magnesium (%)	0202	0.19350	0.01900	0.21593	0.01136	0.00696	99	-1.97	027.43	0
027	Magnesium (%)	0870	0.19435	0.00410	0.21593	0.01136	0.00696	99	-1.90	027.42	0
027	Magnesium (%)	0861	0.19750	0.00100	0.21593	0.01136	0.00696	99	-1.62	027.43	0
027	Magnesium (%)	0939	0.20000	0.00000	0.21593	0.01136	0.00696	99	-1.40	027.31	0
027	Magnesium (%)	0242	0.20000	0.00000	0.21593	0.01136	0.00696	99	-1.40	027.99	0
027	Magnesium (%)	0042	0.20350	0.00300	0.21593	0.01136	0.00696	99	-1.09	027.42	0
027	Magnesium (%)	0889	0.20490	0.02740	0.21593	0.01136	0.00696	99	-0.97	027.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
027	Magnesium (%)	0689	0.20500	0.01000	0.21593	0.01136	0.00696	99	-0.96	027.31	0
027	Magnesium (%)	0049	0.20500	0.01000	0.21593	0.01136	0.00696	99	-0.96	027.41	0
027	Magnesium (%)	0520	0.20500	0.03000	0.21593	0.01136	0.00696	99	-0.96	027.41	0
027	Magnesium (%)	0910	0.20500	0.01000	0.21593	0.01136	0.00696	99	-0.96	027.41	0
027	Magnesium (%)	0098	0.20500	0.01000	0.21593	0.01136	0.00696	99	-0.96	027.43	0
027	Magnesium (%)	0675	0.20500	0.03000	0.21593	0.01136	0.00696	99	-0.96	027.43	0
027	Magnesium (%)	0263	0.20760	0.00080	0.21593	0.01136	0.00696	99	-0.73	027.42	0
027	Magnesium (%)	0142	0.20875	0.00310	0.21593	0.01136	0.00696	99	-0.63	027.31	0
027	Magnesium (%)	0171	0.20900	0.00200	0.21593	0.01136	0.00696	99	-0.61	027.41	0
027	Magnesium (%)	0175	0.21000	0.00000	0.21593	0.01136	0.00696	99	-0.52	027.31	0
027	Magnesium (%)	0884	0.21000	0.00000	0.21593	0.01136	0.00696	99	-0.52	027.31	0
027	Magnesium (%)	0169	0.21000	0.00000	0.21593	0.01136	0.00696	99	-0.52	027.32	0
027	Magnesium (%)	0229	0.21000	0.00000	0.21593	0.01136	0.00696	99	-0.52	027.41	0
027	Magnesium (%)	0413	0.21000	0.00000	0.21593	0.01136	0.00696	99	-0.52	027.41	0
027	Magnesium (%)	0511	0.21000	0.00000	0.21593	0.01136	0.00696	99	-0.52	027.41	0
027	Magnesium (%)	0278	0.21000	0.00000	0.21593	0.01136	0.00696	99	-0.52	027.42	0
027	Magnesium (%)	0035	0.21000	0.00000	0.21593	0.01136	0.00696	99	-0.52	027.43	0
027	Magnesium (%)	0610	0.21000	0.00000	0.21593	0.01136	0.00696	99	-0.52	027.43	0
027	Magnesium (%)	0968	0.21000	0.00000	0.21593	0.01136	0.00696	99	-0.52	027.43	0
027	Magnesium (%)	2089	0.21000	0.00000	0.21593	0.01136	0.00696	99	-0.52	027.43	0
027	Magnesium (%)	0590	0.21000	0.00000	0.21593	0.01136	0.00696	99	-0.52	027.99	0
027	Magnesium (%)	2161	0.21000	0.00000	0.21593	0.01136	0.00696	99	-0.52	027.99	0
027	Magnesium (%)	0148	0.21090	0.00000	0.21593	0.01136	0.00696	99	-0.44	027.41	0
027	Magnesium (%)	0045	0.21100	0.01400	0.21593	0.01136	0.00696	99	-0.43	027.42	0
027	Magnesium (%)	0043	0.21100	0.00200	0.21593	0.01136	0.00696	99	-0.43	027.43	0
027	Magnesium (%)	0186	0.21100	0.00000	0.21593	0.01136	0.00696	99	-0.43	027.52	0
027	Magnesium (%)	0034	0.21200	0.00700	0.21593	0.01136	0.00696	99	-0.35	027.43	0
027	Magnesium (%)	0553	0.21200	0.00400	0.21593	0.01136	0.00696	99	-0.35	027.53	0
027	Magnesium (%)	0164	0.21200	0.01800	0.21593	0.01136	0.00696	99	-0.35	027.41	0
027	Magnesium (%)	0504	0.21210	0.00360	0.21593	0.01136	0.00696	99	-0.34	027.32	0
027	Magnesium (%)	0038	0.21415	0.01910	0.21593	0.01136	0.00696	99	-0.16	027.31	0
027	Magnesium (%)	0536	0.21500	0.01000	0.21593	0.01136	0.00696	99	-0.08	027.31	0
027	Magnesium (%)	0629	0.21500	0.01000	0.21593	0.01136	0.00696	99	-0.08	027.31	0
027	Magnesium (%)	0868	0.21500	0.01000	0.21593	0.01136	0.00696	99	-0.08	027.31	0
027	Magnesium (%)	2022	0.21500	0.00000	0.21593	0.01136	0.00696	99	-0.08	027.31	0
027	Magnesium (%)	0035	0.21500	0.01000	0.21593	0.01136	0.00696	99	-0.08	027.32	0
027	Magnesium (%)	0003	0.21500	0.01000	0.21593	0.01136	0.00696	99	-0.08	027.41	0
027	Magnesium (%)	0098	0.21500	0.01000	0.21593	0.01136	0.00696	99	-0.08	027.41	0
027	Magnesium (%)	0035	0.21500	0.01000	0.21593	0.01136	0.00696	99	-0.08	027.42	0
027	Magnesium (%)	0190	0.21500	0.01000	0.21593	0.01136	0.00696	99	-0.08	027.42	0
027	Magnesium (%)	0685	0.21500	0.01000	0.21593	0.01136	0.00696	99	-0.08	027.42	0
027	Magnesium (%)	0083	0.21500	0.01000	0.21593	0.01136	0.00696	99	-0.08	027.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
027	Magnesium (%)	0297	0.21500	0.01000	0.21593	0.01136	0.00696	99	-0.08	027.43	0
027	Magnesium (%)	0510	0.21500	0.01000	0.21593	0.01136	0.00696	99	-0.08	027.43	0
027	Magnesium (%)	0265	0.21500	0.03000	0.21593	0.01136	0.00696	99	-0.08	027.42	0
027	Magnesium (%)	0874	0.21700	0.00400	0.21593	0.01136	0.00696	99	0.09	027.31	0
027	Magnesium (%)	0186	0.21800	0.00000	0.21593	0.01136	0.00696	99	0.18	027.42	0
027	Magnesium (%)	0918	0.21800	0.00400	0.21593	0.01136	0.00696	99	0.18	027.43	0
027	Magnesium (%)	0560	0.21805	0.02730	0.21593	0.01136	0.00696	99	0.19	027.42	0
027	Magnesium (%)	0017	0.21850	0.01300	0.21593	0.01136	0.00696	99	0.23	027.43	0
027	Magnesium (%)	0139	0.21880	0.00300	0.21593	0.01136	0.00696	99	0.25	027.41	0
027	Magnesium (%)	0964	0.21960	0.01020	0.21593	0.01136	0.00696	99	0.32	027.43	0
027	Magnesium (%)	0354	0.22000	0.00000	0.21593	0.01136	0.00696	99	0.36	027.31	0
027	Magnesium (%)	2196	0.22000	0.00000	0.21593	0.01136	0.00696	99	0.36	027.31	0
027	Magnesium (%)	0019	0.22000	0.00000	0.21593	0.01136	0.00696	99	0.36	027.41	0
027	Magnesium (%)	0074	0.22000	0.00000	0.21593	0.01136	0.00696	99	0.36	027.41	0
027	Magnesium (%)	0878	0.22000	0.02000	0.21593	0.01136	0.00696	99	0.36	027.41	0
027	Magnesium (%)	2053	0.22000	0.00000	0.21593	0.01136	0.00696	99	0.36	027.42	0
027	Magnesium (%)	0357	0.22010	0.00160	0.21593	0.01136	0.00696	99	0.37	027.42	0
027	Magnesium (%)	0208	0.22150	0.00300	0.21593	0.01136	0.00696	99	0.49	027.41	0
027	Magnesium (%)	0598	0.22215	0.00250	0.21593	0.01136	0.00696	99	0.55	027.41	0
027	Magnesium (%)	0208	0.22300	0.01000	0.21593	0.01136	0.00696	99	0.62	027.31	0
027	Magnesium (%)	0726	0.22465	0.00010	0.21593	0.01136	0.00696	99	0.77	027.42	0
027	Magnesium (%)	0037	0.22500	0.01000	0.21593	0.01136	0.00696	99	0.80	027.42	0
027	Magnesium (%)	0294	0.22500	0.01000	0.21593	0.01136	0.00696	99	0.80	027.42	0
027	Magnesium (%)	0425	0.22500	0.01000	0.21593	0.01136	0.00696	99	0.80	027.43	0
027	Magnesium (%)	0032	0.22525	0.01250	0.21593	0.01136	0.00696	99	0.82	027.43	0
027	Magnesium (%)	0187	0.22655	0.00050	0.21593	0.01136	0.00696	99	0.93	027.42	0
027	Magnesium (%)	0038	0.22700	0.00200	0.21593	0.01136	0.00696	99	0.97	027.43	0
027	Magnesium (%)	0405	0.22800	0.00400	0.21593	0.01136	0.00696	99	1.06	027.41	0
027	Magnesium (%)	0572	0.22850	0.01500	0.21593	0.01136	0.00696	99	1.11	027.53	0
027	Magnesium (%)	0653	0.22950	0.00300	0.21593	0.01136	0.00696	99	1.19	027.41	0
027	Magnesium (%)	0629	0.23000	0.02000	0.21593	0.01136	0.00696	99	1.24	027.43	0
027	Magnesium (%)	0226	0.23000	0.00000	0.21593	0.01136	0.00696	99	1.24	027.41	0
027	Magnesium (%)	0505	0.23000	0.00400	0.21593	0.01136	0.00696	99	1.24	027.41	0
027	Magnesium (%)	0720	0.23000	0.00000	0.21593	0.01136	0.00696	99	1.24	027.41	0
027	Magnesium (%)	0693	0.23000	0.00000	0.21593	0.01136	0.00696	99	1.24	027.42	0
027	Magnesium (%)	0100	0.23000	0.00000	0.21593	0.01136	0.00696	99	1.24	027.99	0
027	Magnesium (%)	0563	0.23605	0.00230	0.21593	0.01136	0.00696	99	1.77	027.31	0
027	Magnesium (%)	0682	0.23810	0.00000	0.21593	0.01136	0.00696	99	1.95	027.43	0
027	Magnesium (%)	0358	0.24000	0.00000	0.21593	0.01136	0.00696	99	2.12	027.41	0
027	Magnesium (%)	0154	0.24210	0.01400	0.21593	0.01136	0.00696	99	2.30	027.52	0
027	Magnesium (%)	0353	0.24500	0.01000	0.21593	0.01136	0.00696	99	2.56	027.43	0
027	Magnesium (%)	0089	0.25000	0.00000	0.21593	0.01136	0.00696	99	3.00	027.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
027	Magnesium (%)	0650	0.26900	0.02000	0.21593	0.01136	0.00696	99	4.67	027.31	0
027	Magnesium (%)	0948	0.27200	0.00600	0.21593	0.01136	0.00696	99	4.94	027.33	0
027	Magnesium (%)	0609	0.31500	0.01000	0.21593	0.01136	0.00696	99	8.72	027.32	0
027	Magnesium (%)	2178	2,332.0	26.000	0.21593	0.01136	0.00696	99	205240.77	027.33	2
028	Manganese (mg / kg (ppm))	0405	118.00	2.0000	181.89	10.805	6.1147	99	-5.91	028.41	0
028	Manganese (mg / kg (ppm))	0169	131.00	8.0000	181.89	10.805	6.1147	99	-4.71	028.43	0
028	Manganese (mg / kg (ppm))	0505	140.05	4.1000	181.89	10.805	6.1147	99	-3.87	028.31	0
028	Manganese (mg / kg (ppm))	0047	142.40	3.8000	181.89	10.805	6.1147	99	-3.65	028.52	0
028	Manganese (mg / kg (ppm))	0049	155.04	14.110	181.89	10.805	6.1147	99	-2.49	028.41	0
028	Manganese (mg / kg (ppm))	0870	156.35	4.5000	181.89	10.805	6.1147	99	-2.36	028.42	0
028	Manganese (mg / kg (ppm))	0609	159.00	12.000	181.89	10.805	6.1147	99	-2.12	028.32	0
028	Manganese (mg / kg (ppm))	0675	160.73	11.060	181.89	10.805	6.1147	99	-1.96	028.43	0
028	Manganese (mg / kg (ppm))	0045	163.50	3.0000	181.89	10.805	6.1147	99	-1.70	028.42	0
028	Manganese (mg / kg (ppm))	0653	167.40	1.6000	181.89	10.805	6.1147	99	-1.34	028.41	0
028	Manganese (mg / kg (ppm))	0685	167.63	1.1800	181.89	10.805	6.1147	99	-1.32	028.42	0
028	Manganese (mg / kg (ppm))	0520	168.00	18.000	181.89	10.805	6.1147	99	-1.29	028.41	0
028	Manganese (mg / kg (ppm))	0511	168.50	1.0000	181.89	10.805	6.1147	99	-1.24	028.41	0
028	Manganese (mg / kg (ppm))	0171	169.80	1.8000	181.89	10.805	6.1147	99	-1.12	028.41	0
028	Manganese (mg / kg (ppm))	0878	170.00	4.0000	181.89	10.805	6.1147	99	-1.10	028.41	0
028	Manganese (mg / kg (ppm))	0407	171.08	0.00000	181.89	10.805	6.1147	99	-1.00	028.41	0
028	Manganese (mg / kg (ppm))	2129	171.60	1.4000	181.89	10.805	6.1147	99	-0.95	028.99	0
028	Manganese (mg / kg (ppm))	0668	172.50	9.0000	181.89	10.805	6.1147	99	-0.87	028.43	0
028	Manganese (mg / kg (ppm))	2192	172.50	16.800	181.89	10.805	6.1147	99	-0.87	028.53	0
028	Manganese (mg / kg (ppm))	0353	173.00	10.000	181.89	10.805	6.1147	99	-0.82	028.43	0
028	Manganese (mg / kg (ppm))	0720	173.62	19.860	181.89	10.805	6.1147	99	-0.77	028.41	0
028	Manganese (mg / kg (ppm))	0861	174.50	3.0000	181.89	10.805	6.1147	99	-0.68	028.43	0
028	Manganese (mg / kg (ppm))	0563	174.60	5.4702	181.89	10.805	6.1147	99	-0.67	028.31	0
028	Manganese (mg / kg (ppm))	0037	174.70	8.4000	181.89	10.805	6.1147	99	-0.67	028.42	0
028	Manganese (mg / kg (ppm))	0590	175.00	4.0000	181.89	10.805	6.1147	99	-0.64	028.99	0
028	Manganese (mg / kg (ppm))	0345	175.50	1.0000	181.89	10.805	6.1147	99	-0.59	028.43	0
028	Manganese (mg / kg (ppm))	0098	175.60	12.600	181.89	10.805	6.1147	99	-0.58	028.41	0
028	Manganese (mg / kg (ppm))	0208	175.65	8.3000	181.89	10.805	6.1147	99	-0.58	028.41	0
028	Manganese (mg / kg (ppm))	0033	176.00	2.0000	181.89	10.805	6.1147	99	-0.55	028.43	0
028	Manganese (mg / kg (ppm))	0148	176.05	2.9000	181.89	10.805	6.1147	99	-0.54	028.41	0
028	Manganese (mg / kg (ppm))	2089	176.06	9.7300	181.89	10.805	6.1147	99	-0.54	028.43	0
028	Manganese (mg / kg (ppm))	0598	176.20	9.0000	181.89	10.805	6.1147	99	-0.53	028.41	0
028	Manganese (mg / kg (ppm))	0629	176.50	1.0000	181.89	10.805	6.1147	99	-0.50	028.31	0
028	Manganese (mg / kg (ppm))	0910	177.00	10.000	181.89	10.805	6.1147	99	-0.45	028.41	0
028	Manganese (mg / kg (ppm))	0689	178.45	4.7000	181.89	10.805	6.1147	99	-0.32	028.31	0
028	Manganese (mg / kg (ppm))	0017	179.40	4.9880	181.89	10.805	6.1147	99	-0.23	028.43	0
028	Manganese (mg / kg (ppm))	0003	179.50	5.0000	181.89	10.805	6.1147	99	-0.22	028.41	0
028	Manganese (mg / kg (ppm))	0164	179.50	3.0000	181.89	10.805	6.1147	99	-0.22	028.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
028	Manganese (mg / kg (ppm))	0692	179.50	15.000	181.89	10.805	6.1147	99	-0.22	028.42	0
028	Manganese (mg / kg (ppm))	0043	179.50	9.0000	181.89	10.805	6.1147	99	-0.22	028.43	0
028	Manganese (mg / kg (ppm))	0968	179.50	1.0000	181.89	10.805	6.1147	99	-0.22	028.43	0
028	Manganese (mg / kg (ppm))	0610	180.50	7.0000	181.89	10.805	6.1147	99	-0.13	028.43	0
028	Manganese (mg / kg (ppm))	0034	180.55	0.70000	181.89	10.805	6.1147	99	-0.12	028.43	0
028	Manganese (mg / kg (ppm))	2128	181.32	14.009	181.89	10.805	6.1147	99	-0.05	028.32	0
028	Manganese (mg / kg (ppm))	0425	181.50	2.8000	181.89	10.805	6.1147	99	-0.04	028.43	0
028	Manganese (mg / kg (ppm))	0019	181.70	1.4000	181.89	10.805	6.1147	99	-0.02	028.41	0
028	Manganese (mg / kg (ppm))	0175	182.00	8.0000	181.89	10.805	6.1147	99	0.01	028.31	0
028	Manganese (mg / kg (ppm))	2195	182.00	4.0000	181.89	10.805	6.1147	99	0.01	028.31	0
028	Manganese (mg / kg (ppm))	0413	182.00	12.000	181.89	10.805	6.1147	99	0.01	028.42	0
028	Manganese (mg / kg (ppm))	2196	182.40	0.00000	181.89	10.805	6.1147	99	0.05	028.31	0
028	Manganese (mg / kg (ppm))	0884	182.45	3.3000	181.89	10.805	6.1147	99	0.05	028.31	0
028	Manganese (mg / kg (ppm))	0297	182.50	9.0000	181.89	10.805	6.1147	99	0.06	028.43	0
028	Manganese (mg / kg (ppm))	0553	182.50	3.0000	181.89	10.805	6.1147	99	0.06	028.53	0
028	Manganese (mg / kg (ppm))	0357	182.54	1.9610	181.89	10.805	6.1147	99	0.06	028.42	0
028	Manganese (mg / kg (ppm))	0939	182.90	1.8900	181.89	10.805	6.1147	99	0.09	028.31	0
028	Manganese (mg / kg (ppm))	0964	183.05	6.0273	181.89	10.805	6.1147	99	0.11	028.43	0
028	Manganese (mg / kg (ppm))	2145	183.20	3.8000	181.89	10.805	6.1147	99	0.12	028.41	0
028	Manganese (mg / kg (ppm))	0187	183.47	0.74000	181.89	10.805	6.1147	99	0.15	028.42	0
028	Manganese (mg / kg (ppm))	2022	183.50	7.0000	181.89	10.805	6.1147	99	0.15	028.31	0
028	Manganese (mg / kg (ppm))	0001	184.15	1.5000	181.89	10.805	6.1147	99	0.21	028.31	0
028	Manganese (mg / kg (ppm))	0693	184.50	3.0000	181.89	10.805	6.1147	99	0.24	028.42	0
028	Manganese (mg / kg (ppm))	0098	184.55	5.7000	181.89	10.805	6.1147	99	0.25	028.43	0
028	Manganese (mg / kg (ppm))	0948	185.28	0.83200	181.89	10.805	6.1147	99	0.31	028.33	0
028	Manganese (mg / kg (ppm))	0038	185.55	5.9000	181.89	10.805	6.1147	99	0.34	028.32	0
028	Manganese (mg / kg (ppm))	0868	185.96	1.2300	181.89	10.805	6.1147	99	0.38	028.31	0
028	Manganese (mg / kg (ppm))	0358	186.63	10.400	181.89	10.805	6.1147	99	0.44	028.41	0
028	Manganese (mg / kg (ppm))	0510	187.00	6.0000	181.89	10.805	6.1147	99	0.47	028.43	0
028	Manganese (mg / kg (ppm))	0278	187.50	3.0000	181.89	10.805	6.1147	99	0.52	028.42	0
028	Manganese (mg / kg (ppm))	2053	187.50	11.000	181.89	10.805	6.1147	99	0.52	028.42	0
028	Manganese (mg / kg (ppm))	0629	187.50	5.0000	181.89	10.805	6.1147	99	0.52	028.43	0
028	Manganese (mg / kg (ppm))	0536	187.58	0.65000	181.89	10.805	6.1147	99	0.53	028.31	0
028	Manganese (mg / kg (ppm))	2188	188.58	3.8360	181.89	10.805	6.1147	99	0.62	028.31	0
028	Manganese (mg / kg (ppm))	0229	189.05	5.3000	181.89	10.805	6.1147	99	0.66	028.41	0
028	Manganese (mg / kg (ppm))	0035	189.50	1.0000	181.89	10.805	6.1147	99	0.70	028.43	0
028	Manganese (mg / kg (ppm))	0354	189.65	0.13000	181.89	10.805	6.1147	99	0.72	028.31	0
028	Manganese (mg / kg (ppm))	0726	189.85	0.70000	181.89	10.805	6.1147	99	0.74	028.42	0
028	Manganese (mg / kg (ppm))	0035	190.00	0.00000	181.89	10.805	6.1147	99	0.75	028.42	0
028	Manganese (mg / kg (ppm))	0265	190.00	18.000	181.89	10.805	6.1147	99	0.75	028.42	0
028	Manganese (mg / kg (ppm))	0208	190.50	3.0000	181.89	10.805	6.1147	99	0.80	028.31	0
028	Manganese (mg / kg (ppm))	0074	190.50	7.0000	181.89	10.805	6.1147	99	0.80	028.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
028	Manganese (mg / kg (ppm))	0226	190.50	11.000	181.89	10.805	6.1147	99	0.80	028.41	0
028	Manganese (mg / kg (ppm))	0190	190.76	1.0800	181.89	10.805	6.1147	99	0.82	028.42	0
028	Manganese (mg / kg (ppm))	0560	191.05	27.100	181.89	10.805	6.1147	99	0.85	028.42	0
028	Manganese (mg / kg (ppm))	0038	191.90	7.8800	181.89	10.805	6.1147	99	0.93	028.43	0
028	Manganese (mg / kg (ppm))	0682	193.74	0.00000	181.89	10.805	6.1147	99	1.10	028.43	0
028	Manganese (mg / kg (ppm))	0186	194.00	0.00000	181.89	10.805	6.1147	99	1.12	028.52	0
028	Manganese (mg / kg (ppm))	0572	194.00	22.000	181.89	10.805	6.1147	99	1.12	028.53	0
028	Manganese (mg / kg (ppm))	0202	194.23	12.360	181.89	10.805	6.1147	99	1.14	028.43	0
028	Manganese (mg / kg (ppm))	0242	195.00	10.000	181.89	10.805	6.1147	99	1.21	028.99	0
028	Manganese (mg / kg (ppm))	0083	195.50	7.0000	181.89	10.805	6.1147	99	1.26	028.43	0
028	Manganese (mg / kg (ppm))	0027	195.93	4.9580	181.89	10.805	6.1147	99	1.30	028.43	0
028	Manganese (mg / kg (ppm))	0186	197.00	2.0000	181.89	10.805	6.1147	99	1.40	028.42	0
028	Manganese (mg / kg (ppm))	2144	199.91	2.9700	181.89	10.805	6.1147	99	1.67	028.99	0
028	Manganese (mg / kg (ppm))	0100	201.50	3.0000	181.89	10.805	6.1147	99	1.81	028.99	0
028	Manganese (mg / kg (ppm))	0032	203.65	14.700	181.89	10.805	6.1147	99	2.01	028.43	0
028	Manganese (mg / kg (ppm))	2161	207.45	1.3000	181.89	10.805	6.1147	99	2.37	028.99	0
028	Manganese (mg / kg (ppm))	0504	212.80	9.8000	181.89	10.805	6.1147	99	2.86	028.32	0
028	Manganese (mg / kg (ppm))	0294	224.61	20.620	181.89	10.805	6.1147	99	3.95	028.42	0
028	Manganese (mg / kg (ppm))	0263	224.64	0.48000	181.89	10.805	6.1147	99	3.96	028.42	0
028	Manganese (mg / kg (ppm))	0042	187.00	30.000	181.89	10.805	6.1147	99	0.47	028.42	1
028	Manganese (mg / kg (ppm))	0918	212.93	31.390	181.89	10.805	6.1147	99	2.87	028.43	1
031	Phosphorus (%)	0852	0.84000	0.02000	1.1362	0.05193	0.03194	144	-5.70	031.99	0
031	Phosphorus (%)	0108	0.92000	0.14000	1.1362	0.05193	0.03194	144	-4.16	031.01	0
031	Phosphorus (%)	0889	0.93250	0.02000	1.1362	0.05193	0.03194	144	-3.92	031.99	0
031	Phosphorus (%)	0278	0.97000	0.02000	1.1362	0.05193	0.03194	144	-3.20	031.42	0
031	Phosphorus (%)	2089	0.99500	0.05000	1.1362	0.05193	0.03194	144	-2.72	031.43	0
031	Phosphorus (%)	0405	1.0150	0.01000	1.1362	0.05193	0.03194	144	-2.33	031.41	0
031	Phosphorus (%)	0870	1.0170	0.00800	1.1362	0.05193	0.03194	144	-2.30	031.42	0
031	Phosphorus (%)	2128	1.0355	0.17700	1.1362	0.05193	0.03194	144	-1.94	031.06	0
031	Phosphorus (%)	2199	1.0400	0.10000	1.1362	0.05193	0.03194	144	-1.85	031.01	0
031	Phosphorus (%)	0345	1.0500	0.02000	1.1362	0.05193	0.03194	144	-1.66	031.43	0
031	Phosphorus (%)	0536	1.0550	0.03000	1.1362	0.05193	0.03194	144	-1.56	031.06	0
031	Phosphorus (%)	0590	1.0600	0.04000	1.1362	0.05193	0.03194	144	-1.47	031.01	0
031	Phosphorus (%)	0035	1.0650	0.05000	1.1362	0.05193	0.03194	144	-1.37	031.43	0
031	Phosphorus (%)	0202	1.0695	0.09500	1.1362	0.05193	0.03194	144	-1.29	031.43	0
031	Phosphorus (%)	0940	1.0700	0.02000	1.1362	0.05193	0.03194	144	-1.28	031.01	0
031	Phosphorus (%)	0650	1.0750	0.13000	1.1362	0.05193	0.03194	144	-1.18	031.01	0
031	Phosphorus (%)	0910	1.0750	0.01000	1.1362	0.05193	0.03194	144	-1.18	031.41	0
031	Phosphorus (%)	0045	1.0750	0.01000	1.1362	0.05193	0.03194	144	-1.18	031.42	0
031	Phosphorus (%)	0504	1.0760	0.07400	1.1362	0.05193	0.03194	144	-1.16	031.42	0
031	Phosphorus (%)	0622	1.0774	0.00750	1.1362	0.05193	0.03194	144	-1.13	031.01	0
031	Phosphorus (%)	0848	1.0800	0.00000	1.1362	0.05193	0.03194	144	-1.08	031.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	0042	1.0800	0.02000	1.1362	0.05193	0.03194	144	-1.08	031.42	0
031	Phosphorus (%)	0918	1.0805	0.06500	1.1362	0.05193	0.03194	144	-1.07	031.43	0
031	Phosphorus (%)	0511	1.0850	0.01000	1.1362	0.05193	0.03194	144	-0.99	031.01	0
031	Phosphorus (%)	0720	1.0850	0.01000	1.1362	0.05193	0.03194	144	-0.99	031.41	0
031	Phosphorus (%)	0043	1.0850	0.05000	1.1362	0.05193	0.03194	144	-0.99	031.43	0
031	Phosphorus (%)	0874	1.0878	0.01340	1.1362	0.05193	0.03194	144	-0.93	031.01	0
031	Phosphorus (%)	0006	1.0880	0.06200	1.1362	0.05193	0.03194	144	-0.93	031.43	0
031	Phosphorus (%)	0187	1.0900	0.00000	1.1362	0.05193	0.03194	144	-0.89	031.42	0
031	Phosphorus (%)	0018	1.0950	0.01000	1.1362	0.05193	0.03194	144	-0.79	031.01	0
031	Phosphorus (%)	2192	1.0950	0.03000	1.1362	0.05193	0.03194	144	-0.79	031.01	0
031	Phosphorus (%)	0144	1.0950	0.01000	1.1362	0.05193	0.03194	144	-0.79	031.41	0
031	Phosphorus (%)	0186	1.0950	0.01000	1.1362	0.05193	0.03194	144	-0.79	031.42	0
031	Phosphorus (%)	0563	1.0983	0.01840	1.1362	0.05193	0.03194	144	-0.73	031.01	0
031	Phosphorus (%)	2146	1.1000	0.08000	1.1362	0.05193	0.03194	144	-0.70	031.01	0
031	Phosphorus (%)	2200	1.1000	0.00000	1.1362	0.05193	0.03194	144	-0.70	031.01	0
031	Phosphorus (%)	0169	1.1000	0.02000	1.1362	0.05193	0.03194	144	-0.70	031.03	0
031	Phosphorus (%)	0049	1.1000	0.04000	1.1362	0.05193	0.03194	144	-0.70	031.41	0
031	Phosphorus (%)	0205	1.1000	0.02000	1.1362	0.05193	0.03194	144	-0.70	031.42	0
031	Phosphorus (%)	0010	1.1000	0.06000	1.1362	0.05193	0.03194	144	-0.70	031.43	0
031	Phosphorus (%)	0027	1.1015	0.05100	1.1362	0.05193	0.03194	144	-0.67	031.43	0
031	Phosphorus (%)	0861	1.1050	0.01000	1.1362	0.05193	0.03194	144	-0.60	031.43	0
031	Phosphorus (%)	0142	1.1057	0.00380	1.1362	0.05193	0.03194	144	-0.59	031.01	0
031	Phosphorus (%)	0265	1.1100	0.12000	1.1362	0.05193	0.03194	144	-0.51	031.42	0
031	Phosphorus (%)	0692	1.1100	0.18000	1.1362	0.05193	0.03194	144	-0.51	031.42	0
031	Phosphorus (%)	0019	1.1100	0.02000	1.1362	0.05193	0.03194	144	-0.51	031.41	0
031	Phosphorus (%)	0298	1.1100	0.02000	1.1362	0.05193	0.03194	144	-0.51	031.41	0
031	Phosphorus (%)	0034	1.1145	0.02300	1.1362	0.05193	0.03194	144	-0.42	031.43	0
031	Phosphorus (%)	0300	1.1150	0.09000	1.1362	0.05193	0.03194	144	-0.41	031.99	0
031	Phosphorus (%)	0171	1.1150	0.01000	1.1362	0.05193	0.03194	144	-0.41	031.41	0
031	Phosphorus (%)	2109	1.1160	0.05600	1.1362	0.05193	0.03194	144	-0.39	031.01	0
031	Phosphorus (%)	0849	1.1175	0.00500	1.1362	0.05193	0.03194	144	-0.36	031.01	0
031	Phosphorus (%)	0560	1.1189	0.11680	1.1362	0.05193	0.03194	144	-0.33	031.42	0
031	Phosphorus (%)	0685	1.1200	0.02000	1.1362	0.05193	0.03194	144	-0.31	031.41	0
031	Phosphorus (%)	0948	1.1245	0.00300	1.1362	0.05193	0.03194	144	-0.23	031.01	0
031	Phosphorus (%)	0035	1.1250	0.03000	1.1362	0.05193	0.03194	144	-0.22	031.01	0
031	Phosphorus (%)	0651	1.1250	0.01400	1.1362	0.05193	0.03194	144	-0.22	031.01	0
031	Phosphorus (%)	2145	1.1250	0.01000	1.1362	0.05193	0.03194	144	-0.22	031.41	0
031	Phosphorus (%)	0609	1.1300	0.00000	1.1362	0.05193	0.03194	144	-0.12	031.01	0
031	Phosphorus (%)	0676	1.1300	0.00000	1.1362	0.05193	0.03194	144	-0.12	031.99	0
031	Phosphorus (%)	0689	1.1350	0.01000	1.1362	0.05193	0.03194	144	-0.02	031.01	0
031	Phosphorus (%)	0878	1.1350	0.01000	1.1362	0.05193	0.03194	144	-0.02	031.01	0
031	Phosphorus (%)	0897	1.1350	0.01000	1.1362	0.05193	0.03194	144	-0.02	031.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	2188	1.1350	0.01000	1.1362	0.05193	0.03194	144	-0.02	031.01	0
031	Phosphorus (%)	2129	1.1350	0.01000	1.1362	0.05193	0.03194	144	-0.02	031.42	0
031	Phosphorus (%)	0939	1.1350	0.03000	1.1362	0.05193	0.03194	144	-0.02	031.01	0
031	Phosphorus (%)	2006	1.1350	0.04000	1.1362	0.05193	0.03194	144	-0.02	031.01	0
031	Phosphorus (%)	0626	1.1390	0.03600	1.1362	0.05193	0.03194	144	0.05	031.01	0
031	Phosphorus (%)	2193	1.1400	0.00000	1.1362	0.05193	0.03194	144	0.07	031.01	0
031	Phosphorus (%)	0505	1.1400	0.02000	1.1362	0.05193	0.03194	144	0.07	031.02	0
031	Phosphorus (%)	0089	1.1400	0.00000	1.1362	0.05193	0.03194	144	0.07	031.43	0
031	Phosphorus (%)	0610	1.1400	0.02000	1.1362	0.05193	0.03194	144	0.07	031.43	0
031	Phosphorus (%)	0242	1.1400	0.00000	1.1362	0.05193	0.03194	144	0.07	031.99	0
031	Phosphorus (%)	0629	1.1400	0.04000	1.1362	0.05193	0.03194	144	0.07	031.01	0
031	Phosphorus (%)	0035	1.1400	0.04000	1.1362	0.05193	0.03194	144	0.07	031.42	0
031	Phosphorus (%)	0294	1.1400	0.06000	1.1362	0.05193	0.03194	144	0.07	031.42	0
031	Phosphorus (%)	2009	1.1419	0.00130	1.1362	0.05193	0.03194	144	0.11	031.01	0
031	Phosphorus (%)	0881	1.1425	0.01100	1.1362	0.05193	0.03194	144	0.12	031.01	0
031	Phosphorus (%)	0098	1.1430	0.00400	1.1362	0.05193	0.03194	144	0.13	031.43	0
031	Phosphorus (%)	0968	1.1440	0.00200	1.1362	0.05193	0.03194	144	0.15	031.43	0
031	Phosphorus (%)	0354	1.1450	0.01000	1.1362	0.05193	0.03194	144	0.17	031.01	0
031	Phosphorus (%)	2195	1.1450	0.05000	1.1362	0.05193	0.03194	144	0.17	031.01	0
031	Phosphorus (%)	2144	1.1450	0.03000	1.1362	0.05193	0.03194	144	0.17	031.99	0
031	Phosphorus (%)	0098	1.1460	0.06800	1.1362	0.05193	0.03194	144	0.19	031.41	0
031	Phosphorus (%)	2022	1.1500	0.00000	1.1362	0.05193	0.03194	144	0.27	031.01	0
031	Phosphorus (%)	0358	1.1500	0.02000	1.1362	0.05193	0.03194	144	0.27	031.41	0
031	Phosphorus (%)	0629	1.1500	0.06000	1.1362	0.05193	0.03194	144	0.27	031.43	0
031	Phosphorus (%)	0668	1.1500	0.02000	1.1362	0.05193	0.03194	144	0.27	031.43	0
031	Phosphorus (%)	0154	1.1530	0.05600	1.1362	0.05193	0.03194	144	0.32	031.52	0
031	Phosphorus (%)	2190	1.1537	0.04010	1.1362	0.05193	0.03194	144	0.34	031.01	0
031	Phosphorus (%)	0726	1.1545	0.00100	1.1362	0.05193	0.03194	144	0.35	031.42	0
031	Phosphorus (%)	0885	1.1547	0.04020	1.1362	0.05193	0.03194	144	0.36	031.01	0
031	Phosphorus (%)	0175	1.1550	0.03000	1.1362	0.05193	0.03194	144	0.36	031.01	0
031	Phosphorus (%)	0683	1.1550	0.01000	1.1362	0.05193	0.03194	144	0.36	031.01	0
031	Phosphorus (%)	0037	1.1550	0.01000	1.1362	0.05193	0.03194	144	0.36	031.42	0
031	Phosphorus (%)	0033	1.1550	0.07000	1.1362	0.05193	0.03194	144	0.36	031.43	0
031	Phosphorus (%)	0148	1.1553	0.01280	1.1362	0.05193	0.03194	144	0.37	031.41	0
031	Phosphorus (%)	0722	1.1590	0.06610	1.1362	0.05193	0.03194	144	0.44	031.01	0
031	Phosphorus (%)	2196	1.1600	0.00000	1.1362	0.05193	0.03194	144	0.46	031.01	0
031	Phosphorus (%)	0190	1.1600	0.04000	1.1362	0.05193	0.03194	144	0.46	031.42	0
031	Phosphorus (%)	0100	1.1600	0.00000	1.1362	0.05193	0.03194	144	0.46	031.99	0
031	Phosphorus (%)	0139	1.1600	0.01400	1.1362	0.05193	0.03194	144	0.46	031.41	0
031	Phosphorus (%)	0653	1.1610	0.01800	1.1362	0.05193	0.03194	144	0.48	031.41	0
031	Phosphorus (%)	0884	1.1650	0.03000	1.1362	0.05193	0.03194	144	0.55	031.01	0
031	Phosphorus (%)	2126	1.1650	0.01000	1.1362	0.05193	0.03194	144	0.55	031.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	0229	1.1650	0.03000	1.1362	0.05193	0.03194	144	0.55	031.41	0
031	Phosphorus (%)	0868	1.1700	0.00000	1.1362	0.05193	0.03194	144	0.65	031.01	0
031	Phosphorus (%)	0413	1.1700	0.08000	1.1362	0.05193	0.03194	144	0.65	031.41	0
031	Phosphorus (%)	0186	1.1700	0.00000	1.1362	0.05193	0.03194	144	0.65	031.52	0
031	Phosphorus (%)	0357	1.1714	0.01110	1.1362	0.05193	0.03194	144	0.68	031.42	0
031	Phosphorus (%)	0263	1.1720	0.00550	1.1362	0.05193	0.03194	144	0.69	031.42	0
031	Phosphorus (%)	0895	1.1750	0.01000	1.1362	0.05193	0.03194	144	0.75	031.00	0
031	Phosphorus (%)	2053	1.1750	0.01000	1.1362	0.05193	0.03194	144	0.75	031.02	0
031	Phosphorus (%)	0017	1.1750	0.01000	1.1362	0.05193	0.03194	144	0.75	031.43	0
031	Phosphorus (%)	0407	1.1788	0.02890	1.1362	0.05193	0.03194	144	0.82	031.41	0
031	Phosphorus (%)	0164	1.1800	0.02000	1.1362	0.05193	0.03194	144	0.84	031.41	0
031	Phosphorus (%)	0682	1.1800	0.00000	1.1362	0.05193	0.03194	144	0.84	031.43	0
031	Phosphorus (%)	0043	1.1830	0.00600	1.1362	0.05193	0.03194	144	0.90	031.02	0
031	Phosphorus (%)	0964	1.1838	0.00880	1.1362	0.05193	0.03194	144	0.92	031.43	0
031	Phosphorus (%)	0297	1.1850	0.11000	1.1362	0.05193	0.03194	144	0.94	031.43	0
031	Phosphorus (%)	0675	1.1850	0.05000	1.1362	0.05193	0.03194	144	0.94	031.43	0
031	Phosphorus (%)	0208	1.1865	0.00500	1.1362	0.05193	0.03194	144	0.97	031.41	0
031	Phosphorus (%)	0003	1.1900	0.02000	1.1362	0.05193	0.03194	144	1.04	031.41	0
031	Phosphorus (%)	0036	1.1909	0.01030	1.1362	0.05193	0.03194	144	1.05	031.43	0
031	Phosphorus (%)	0038	1.1945	0.10100	1.1362	0.05193	0.03194	144	1.12	031.01	0
031	Phosphorus (%)	0036	1.1948	0.01090	1.1362	0.05193	0.03194	144	1.13	031.03	0
031	Phosphorus (%)	0014	1.1950	0.03000	1.1362	0.05193	0.03194	144	1.13	031.42	0
031	Phosphorus (%)	0553	1.1950	0.05000	1.1362	0.05193	0.03194	144	1.13	031.53	0
031	Phosphorus (%)	0208	1.1960	0.03400	1.1362	0.05193	0.03194	144	1.15	031.01	0
031	Phosphorus (%)	0152	1.2000	0.02000	1.1362	0.05193	0.03194	144	1.23	031.01	0
031	Phosphorus (%)	2181	1.2000	0.04000	1.1362	0.05193	0.03194	144	1.23	031.01	0
031	Phosphorus (%)	0353	1.2050	0.01000	1.1362	0.05193	0.03194	144	1.32	031.43	0
031	Phosphorus (%)	0687	1.2100	0.08000	1.1362	0.05193	0.03194	144	1.42	031.01	0
031	Phosphorus (%)	0226	1.2100	0.02000	1.1362	0.05193	0.03194	144	1.42	031.41	0
031	Phosphorus (%)	0598	1.2230	0.01800	1.1362	0.05193	0.03194	144	1.67	031.41	0
031	Phosphorus (%)	0504	1.2250	0.05000	1.1362	0.05193	0.03194	144	1.71	031.03	0
031	Phosphorus (%)	0083	1.2300	0.02000	1.1362	0.05193	0.03194	144	1.81	031.43	0
031	Phosphorus (%)	0693	1.2350	0.05000	1.1362	0.05193	0.03194	144	1.90	031.42	0
031	Phosphorus (%)	0510	1.2400	0.00000	1.1362	0.05193	0.03194	144	2.00	031.43	0
031	Phosphorus (%)	0074	1.2450	0.01000	1.1362	0.05193	0.03194	144	2.09	031.41	0
031	Phosphorus (%)	0425	1.2450	0.01000	1.1362	0.05193	0.03194	144	2.09	031.43	0
031	Phosphorus (%)	0038	1.2900	0.08000	1.1362	0.05193	0.03194	144	2.96	031.43	0
031	Phosphorus (%)	0572	1.3650	0.03000	1.1362	0.05193	0.03194	144	4.40	031.53	0
031	Phosphorus (%)	0032	1.4000	0.11600	1.1362	0.05193	0.03194	144	5.08	031.43	0
031	Phosphorus (%)	0520	0.83000	0.50000	1.1362	0.05193	0.03194	144	-5.90	031.41	1
031	Phosphorus (%)	2161	0.47500	0.01000	1.1362	0.05193	0.03194	144	-12.73	031.01	2
032	Potassium (%)	2199	0.57000	0.00000	0.92175	0.04640	0.02186	100	-7.58	032.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
032	Potassium (%)	0536	0.81000	0.02000	0.92175	0.04640	0.02186	100	-2.41	032.31	0
032	Potassium (%)	0003	0.83000	0.10000	0.92175	0.04640	0.02186	100	-1.98	032.41	0
032	Potassium (%)	2022	0.83200	0.01000	0.92175	0.04640	0.02186	100	-1.93	032.31	0
032	Potassium (%)	0874	0.85000	0.02400	0.92175	0.04640	0.02186	100	-1.55	032.31	0
032	Potassium (%)	0035	0.85000	0.02000	0.92175	0.04640	0.02186	100	-1.55	032.43	0
032	Potassium (%)	0034	0.85305	0.01050	0.92175	0.04640	0.02186	100	-1.48	032.43	0
032	Potassium (%)	0045	0.85650	0.00100	0.92175	0.04640	0.02186	100	-1.41	032.42	0
032	Potassium (%)	0629	0.86000	0.02000	0.92175	0.04640	0.02186	100	-1.33	032.31	0
032	Potassium (%)	2146	0.86250	0.03300	0.92175	0.04640	0.02186	100	-1.28	032.31	0
032	Potassium (%)	2129	0.86355	0.06650	0.92175	0.04640	0.02186	100	-1.25	032.42	0
032	Potassium (%)	0590	0.86500	0.01000	0.92175	0.04640	0.02186	100	-1.22	032.99	0
032	Potassium (%)	0169	0.87000	0.00000	0.92175	0.04640	0.02186	100	-1.12	032.32	0
032	Potassium (%)	0019	0.87000	0.02000	0.92175	0.04640	0.02186	100	-1.12	032.41	0
032	Potassium (%)	0265	0.87000	0.08000	0.92175	0.04640	0.02186	100	-1.12	032.42	0
032	Potassium (%)	0870	0.87335	0.00550	0.92175	0.04640	0.02186	100	-1.04	032.42	0
032	Potassium (%)	0720	0.87500	0.01000	0.92175	0.04640	0.02186	100	-1.01	032.41	0
032	Potassium (%)	0653	0.87750	0.01100	0.92175	0.04640	0.02186	100	-0.95	032.41	0
032	Potassium (%)	0175	0.88500	0.01000	0.92175	0.04640	0.02186	100	-0.79	032.31	0
032	Potassium (%)	0144	0.88500	0.03000	0.92175	0.04640	0.02186	100	-0.79	032.41	0
032	Potassium (%)	0229	0.89000	0.02000	0.92175	0.04640	0.02186	100	-0.68	032.41	0
032	Potassium (%)	0511	0.89000	0.02000	0.92175	0.04640	0.02186	100	-0.68	032.41	0
032	Potassium (%)	0278	0.89000	0.02000	0.92175	0.04640	0.02186	100	-0.68	032.42	0
032	Potassium (%)	0868	0.89500	0.01000	0.92175	0.04640	0.02186	100	-0.58	032.31	0
032	Potassium (%)	0049	0.89500	0.01000	0.92175	0.04640	0.02186	100	-0.58	032.41	0
032	Potassium (%)	0910	0.89500	0.05000	0.92175	0.04640	0.02186	100	-0.58	032.41	0
032	Potassium (%)	0035	0.89500	0.01000	0.92175	0.04640	0.02186	100	-0.58	032.42	0
032	Potassium (%)	0098	0.89500	0.03000	0.92175	0.04640	0.02186	100	-0.58	032.43	0
032	Potassium (%)	0353	0.89500	0.03000	0.92175	0.04640	0.02186	100	-0.58	032.43	0
032	Potassium (%)	0038	0.89870	0.02900	0.92175	0.04640	0.02186	100	-0.50	032.31	0
032	Potassium (%)	0881	0.89950	0.00900	0.92175	0.04640	0.02186	100	-0.48	032.99	0
032	Potassium (%)	0878	0.90000	0.00000	0.92175	0.04640	0.02186	100	-0.47	032.41	0
032	Potassium (%)	0043	0.90450	0.00300	0.92175	0.04640	0.02186	100	-0.37	032.43	0
032	Potassium (%)	0861	0.90450	0.03700	0.92175	0.04640	0.02186	100	-0.37	032.43	0
032	Potassium (%)	0650	0.90500	0.03000	0.92175	0.04640	0.02186	100	-0.36	032.31	0
032	Potassium (%)	0405	0.90500	0.01000	0.92175	0.04640	0.02186	100	-0.36	032.41	0
032	Potassium (%)	0345	0.90800	0.00400	0.92175	0.04640	0.02186	100	-0.30	032.43	0
032	Potassium (%)	0208	0.90850	0.00100	0.92175	0.04640	0.02186	100	-0.29	032.41	0
032	Potassium (%)	0553	0.90950	0.00900	0.92175	0.04640	0.02186	100	-0.26	032.53	0
032	Potassium (%)	2192	0.91000	0.06000	0.92175	0.04640	0.02186	100	-0.25	032.43	0
032	Potassium (%)	0505	0.91000	0.02000	0.92175	0.04640	0.02186	100	-0.25	032.41	0
032	Potassium (%)	2089	0.91000	0.02000	0.92175	0.04640	0.02186	100	-0.25	032.43	0
032	Potassium (%)	0042	0.91100	0.00800	0.92175	0.04640	0.02186	100	-0.23	032.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
032	Potassium (%)	0668	0.91350	0.00500	0.92175	0.04640	0.02186	100	-0.18	032.43	0
032	Potassium (%)	0629	0.91500	0.01000	0.92175	0.04640	0.02186	100	-0.15	032.43	0
032	Potassium (%)	0918	0.91550	0.02300	0.92175	0.04640	0.02186	100	-0.13	032.43	0
032	Potassium (%)	0504	0.92000	0.04000	0.92175	0.04640	0.02186	100	-0.04	032.02	0
032	Potassium (%)	0100	0.92000	0.04000	0.92175	0.04640	0.02186	100	-0.04	032.99	0
032	Potassium (%)	0354	0.92000	0.02000	0.92175	0.04640	0.02186	100	-0.04	032.31	0
032	Potassium (%)	0939	0.92000	0.00000	0.92175	0.04640	0.02186	100	-0.04	032.31	0
032	Potassium (%)	0074	0.92000	0.00000	0.92175	0.04640	0.02186	100	-0.04	032.41	0
032	Potassium (%)	0682	0.92000	0.00000	0.92175	0.04640	0.02186	100	-0.04	032.43	0
032	Potassium (%)	0263	0.92085	0.00390	0.92175	0.04640	0.02186	100	-0.02	032.42	0
032	Potassium (%)	0504	0.92100	0.04000	0.92175	0.04640	0.02186	100	-0.02	032.42	0
032	Potassium (%)	0164	0.92250	0.03300	0.92175	0.04640	0.02186	100	0.02	032.41	0
032	Potassium (%)	0037	0.92500	0.03000	0.92175	0.04640	0.02186	100	0.07	032.42	0
032	Potassium (%)	0693	0.92500	0.01000	0.92175	0.04640	0.02186	100	0.07	032.42	0
032	Potassium (%)	0964	0.92815	0.04190	0.92175	0.04640	0.02186	100	0.14	032.43	0
032	Potassium (%)	0148	0.92945	0.01670	0.92175	0.04640	0.02186	100	0.17	032.41	0
032	Potassium (%)	0520	0.93000	0.08000	0.92175	0.04640	0.02186	100	0.18	032.41	0
032	Potassium (%)	0294	0.93000	0.02000	0.92175	0.04640	0.02186	100	0.18	032.42	0
032	Potassium (%)	0610	0.93000	0.02000	0.92175	0.04640	0.02186	100	0.18	032.43	0
032	Potassium (%)	0242	0.93000	0.00000	0.92175	0.04640	0.02186	100	0.18	032.99	0
032	Potassium (%)	0186	0.93450	0.00300	0.92175	0.04640	0.02186	100	0.27	032.42	0
032	Potassium (%)	0171	0.93500	0.01000	0.92175	0.04640	0.02186	100	0.29	032.41	0
032	Potassium (%)	0208	0.93600	0.01200	0.92175	0.04640	0.02186	100	0.31	032.31	0
032	Potassium (%)	0205	0.94000	0.00000	0.92175	0.04640	0.02186	100	0.39	032.42	0
032	Potassium (%)	0425	0.94000	0.02000	0.92175	0.04640	0.02186	100	0.39	032.43	0
032	Potassium (%)	0186	0.94100	0.01600	0.92175	0.04640	0.02186	100	0.41	032.52	0
032	Potassium (%)	0187	0.94130	0.00280	0.92175	0.04640	0.02186	100	0.42	032.42	0
032	Potassium (%)	2053	0.94500	0.01000	0.92175	0.04640	0.02186	100	0.50	032.31	0
032	Potassium (%)	0563	0.94880	0.00440	0.92175	0.04640	0.02186	100	0.58	032.31	0
032	Potassium (%)	0968	0.94950	0.00100	0.92175	0.04640	0.02186	100	0.60	032.43	0
032	Potassium (%)	0358	0.95000	0.02000	0.92175	0.04640	0.02186	100	0.61	032.41	0
032	Potassium (%)	0675	0.95000	0.02000	0.92175	0.04640	0.02186	100	0.61	032.43	0
032	Potassium (%)	0098	0.95500	0.05000	0.92175	0.04640	0.02186	100	0.72	032.41	0
032	Potassium (%)	0032	0.95570	0.03300	0.92175	0.04640	0.02186	100	0.73	032.43	0
032	Potassium (%)	0560	0.95590	0.11080	0.92175	0.04640	0.02186	100	0.74	032.42	0
032	Potassium (%)	0139	0.95700	0.00000	0.92175	0.04640	0.02186	100	0.76	032.41	0
032	Potassium (%)	0038	0.95750	0.00500	0.92175	0.04640	0.02186	100	0.77	032.43	0
032	Potassium (%)	0017	0.96100	0.05200	0.92175	0.04640	0.02186	100	0.85	032.43	0
032	Potassium (%)	0884	0.96500	0.01000	0.92175	0.04640	0.02186	100	0.93	032.02	0
032	Potassium (%)	0226	0.96500	0.01000	0.92175	0.04640	0.02186	100	0.93	032.41	0
032	Potassium (%)	0413	0.96500	0.01000	0.92175	0.04640	0.02186	100	0.93	032.41	0
032	Potassium (%)	0598	0.96545	0.00230	0.92175	0.04640	0.02186	100	0.94	032.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
032	Potassium (%)	0683	0.97000	0.02000	0.92175	0.04640	0.02186	100	1.04	032.31	0
032	Potassium (%)	0190	0.97000	0.02000	0.92175	0.04640	0.02186	100	1.04	032.42	0
032	Potassium (%)	0948	0.97800	0.01400	0.92175	0.04640	0.02186	100	1.21	032.31	0
032	Potassium (%)	0407	0.98250	0.00660	0.92175	0.04640	0.02186	100	1.31	032.41	0
032	Potassium (%)	0726	0.98805	0.00150	0.92175	0.04640	0.02186	100	1.43	032.42	0
032	Potassium (%)	0083	0.99500	0.03000	0.92175	0.04640	0.02186	100	1.58	032.43	0
032	Potassium (%)	0297	0.99500	0.01000	0.92175	0.04640	0.02186	100	1.58	032.43	0
032	Potassium (%)	0142	0.99930	0.00720	0.92175	0.04640	0.02186	100	1.67	032.31	0
032	Potassium (%)	0357	0.99930	0.01900	0.92175	0.04640	0.02186	100	1.67	032.42	0
032	Potassium (%)	0609	1.0050	0.01000	0.92175	0.04640	0.02186	100	1.79	032.32	0
032	Potassium (%)	0089	1.0100	0.00000	0.92175	0.04640	0.02186	100	1.90	032.43	0
032	Potassium (%)	0510	1.0100	0.02000	0.92175	0.04640	0.02186	100	1.90	032.43	0
032	Potassium (%)	0154	1.0272	0.05070	0.92175	0.04640	0.02186	100	2.27	032.52	0
032	Potassium (%)	0572	1.2100	0.12000	0.92175	0.04640	0.02186	100	6.21	032.53	0
032	Potassium (%)	2161	1.2500	0.10000	0.92175	0.04640	0.02186	100	7.07	032.99	0
032	Potassium (%)	0202	0.85050	0.14900	0.92175	0.04640	0.02186	100	-1.54	032.43	1
032	Potassium (%)	0692	0.92000	0.14000	0.92175	0.04640	0.02186	100	-0.04	032.42	1
032	Potassium (%)	0889	0.21130	0.06740	0.92175	0.04640	0.02186	100	-15.31	032.99	2
033	Salt as chloride (%)	2129	1.1250	0.09000	2.0065	0.10089	0.02938	73	-8.74	033.05	0
033	Salt as chloride (%)	0536	1.2950	0.09000	2.0065	0.10089	0.02938	73	-7.05	033.99	0
033	Salt as chloride (%)	0505	1.7000	0.06000	2.0065	0.10089	0.02938	73	-3.04	033.03	0
033	Salt as chloride (%)	0881	1.7575	0.03300	2.0065	0.10089	0.02938	73	-2.47	033.00	0
033	Salt as chloride (%)	0884	1.8050	0.03000	2.0065	0.10089	0.02938	73	-2.00	033.00	0
033	Salt as chloride (%)	2196	1.8100	0.00000	2.0065	0.10089	0.02938	73	-1.95	033.00	0
033	Salt as chloride (%)	0298	1.8400	0.02000	2.0065	0.10089	0.02938	73	-1.65	033.00	0
033	Salt as chloride (%)	0868	1.8450	0.01000	2.0065	0.10089	0.02938	73	-1.60	033.05	0
033	Salt as chloride (%)	0171	1.8500	0.00000	2.0065	0.10089	0.02938	73	-1.55	033.00	0
033	Salt as chloride (%)	0017	1.8995	0.00500	2.0065	0.10089	0.02938	73	-1.06	033.99	0
033	Salt as chloride (%)	2076	1.9100	0.00000	2.0065	0.10089	0.02938	73	-0.96	033.00	0
033	Salt as chloride (%)	0511	1.9150	0.09000	2.0065	0.10089	0.02938	73	-0.91	033.00	0
033	Salt as chloride (%)	0878	1.9150	0.01000	2.0065	0.10089	0.02938	73	-0.91	033.00	0
033	Salt as chloride (%)	0353	1.9200	0.04000	2.0065	0.10089	0.02938	73	-0.86	033.00	0
033	Salt as chloride (%)	0895	1.9250	0.01000	2.0065	0.10089	0.02938	73	-0.81	033.00	0
033	Salt as chloride (%)	0358	1.9250	0.01000	2.0065	0.10089	0.02938	73	-0.81	033.99	0
033	Salt as chloride (%)	0190	1.9300	0.14000	2.0065	0.10089	0.02938	73	-0.76	033.03	0
033	Salt as chloride (%)	2181	1.9400	0.04000	2.0065	0.10089	0.02938	73	-0.66	033.99	0
033	Salt as chloride (%)	2145	1.9450	0.01000	2.0065	0.10089	0.02938	73	-0.61	033.00	0
033	Salt as chloride (%)	0169	1.9500	0.00000	2.0065	0.10089	0.02938	73	-0.56	033.00	0
033	Salt as chloride (%)	0208	1.9500	0.04000	2.0065	0.10089	0.02938	73	-0.56	033.99	0
033	Salt as chloride (%)	2009	1.9509	0.01110	2.0065	0.10089	0.02938	73	-0.55	033.00	0
033	Salt as chloride (%)	0229	1.9550	0.01000	2.0065	0.10089	0.02938	73	-0.51	033.01	0
033	Salt as chloride (%)	2022	1.9575	0.00700	2.0065	0.10089	0.02938	73	-0.49	033.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
033	Salt as chloride (%)	0726	1.9750	0.07000	2.0065	0.10089	0.02938	73	-0.31	033.03	0
033	Salt as chloride (%)	0045	1.9750	0.03000	2.0065	0.10089	0.02938	73	-0.31	033.00	0
033	Salt as chloride (%)	0940	1.9750	0.01000	2.0065	0.10089	0.02938	73	-0.31	033.99	0
033	Salt as chloride (%)	0964	1.9780	0.04600	2.0065	0.10089	0.02938	73	-0.28	033.01	0
033	Salt as chloride (%)	0693	1.9850	0.05000	2.0065	0.10089	0.02938	73	-0.21	033.00	0
033	Salt as chloride (%)	0897	1.9850	0.05000	2.0065	0.10089	0.02938	73	-0.21	033.00	0
033	Salt as chloride (%)	0939	1.9950	0.05000	2.0065	0.10089	0.02938	73	-0.11	033.01	0
033	Salt as chloride (%)	2109	1.9985	0.01300	2.0065	0.10089	0.02938	73	-0.08	033.01	0
033	Salt as chloride (%)	0027	2.0055	0.08700	2.0065	0.10089	0.02938	73	-0.01	033.99	0
033	Salt as chloride (%)	0948	2.0060	0.00000	2.0065	0.10089	0.02938	73	0.00	033.01	0
033	Salt as chloride (%)	0083	2.0150	0.01000	2.0065	0.10089	0.02938	73	0.08	033.01	0
033	Salt as chloride (%)	0098	2.0150	0.01000	2.0065	0.10089	0.02938	73	0.08	033.01	0
033	Salt as chloride (%)	0226	2.0200	0.02000	2.0065	0.10089	0.02938	73	0.13	033.01	0
033	Salt as chloride (%)	0683	2.0200	0.02000	2.0065	0.10089	0.02938	73	0.13	033.01	0
033	Salt as chloride (%)	2146	2.0200	0.02000	2.0065	0.10089	0.02938	73	0.13	033.01	0
033	Salt as chloride (%)	0510	2.0200	0.04000	2.0065	0.10089	0.02938	73	0.13	033.01	0
033	Salt as chloride (%)	0861	2.0200	0.00000	2.0065	0.10089	0.02938	73	0.13	033.01	0
033	Salt as chloride (%)	0689	2.0200	0.04000	2.0065	0.10089	0.02938	73	0.13	033.05	0
033	Salt as chloride (%)	0407	2.0245	0.00100	2.0065	0.10089	0.02938	73	0.18	033.01	0
033	Salt as chloride (%)	0650	2.0250	0.01000	2.0065	0.10089	0.02938	73	0.18	033.01	0
033	Salt as chloride (%)	0590	2.0300	0.08000	2.0065	0.10089	0.02938	73	0.23	033.01	0
033	Salt as chloride (%)	0885	2.0354	0.00650	2.0065	0.10089	0.02938	73	0.29	033.00	0
033	Salt as chloride (%)	0001	2.0400	0.00000	2.0065	0.10089	0.02938	73	0.33	033.01	0
033	Salt as chloride (%)	0242	2.0550	0.01000	2.0065	0.10089	0.02938	73	0.48	033.01	0
033	Salt as chloride (%)	0278	2.0550	0.01000	2.0065	0.10089	0.02938	73	0.48	033.01	0
033	Salt as chloride (%)	0148	2.0550	0.00000	2.0065	0.10089	0.02938	73	0.48	033.01	0
033	Salt as chloride (%)	0629	2.0600	0.02000	2.0065	0.10089	0.02938	73	0.53	033.01	0
033	Salt as chloride (%)	0644	2.0623	0.05900	2.0065	0.10089	0.02938	73	0.55	033.01	0
033	Salt as chloride (%)	0504	2.0650	0.01000	2.0065	0.10089	0.02938	73	0.58	033.00	0
033	Salt as chloride (%)	0309	2.0657	0.04020	2.0065	0.10089	0.02938	73	0.59	033.00	0
033	Salt as chloride (%)	0874	2.0700	0.01600	2.0065	0.10089	0.02938	73	0.63	033.01	0
033	Salt as chloride (%)	2073	2.0700	0.00000	2.0065	0.10089	0.02938	73	0.63	033.99	0
033	Salt as chloride (%)	0100	2.0700	0.02000	2.0065	0.10089	0.02938	73	0.63	033.01	0
033	Salt as chloride (%)	0354	2.0800	0.00000	2.0065	0.10089	0.02938	73	0.73	033.01	0
033	Salt as chloride (%)	0413	2.0800	0.04000	2.0065	0.10089	0.02938	73	0.73	033.01	0
033	Salt as chloride (%)	0205	2.0850	0.07000	2.0065	0.10089	0.02938	73	0.78	033.01	0
033	Salt as chloride (%)	0164	2.0900	0.02000	2.0065	0.10089	0.02938	73	0.83	033.01	0
033	Salt as chloride (%)	0175	2.0900	0.02000	2.0065	0.10089	0.02938	73	0.83	033.01	0
033	Salt as chloride (%)	2161	2.0950	0.05000	2.0065	0.10089	0.02938	73	0.88	033.05	0
033	Salt as chloride (%)	0019	2.1150	0.05000	2.0065	0.10089	0.02938	73	1.08	033.01	0
033	Salt as chloride (%)	0425	2.1250	0.01000	2.0065	0.10089	0.02938	73	1.17	033.01	0
033	Salt as chloride (%)	2190	2.1265	0.11010	2.0065	0.10089	0.02938	73	1.19	033.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
033	Salt as chloride (%)	2144	2.1450	0.01000	2.0065	0.10089	0.02938	73	1.37	033.99	0
033	Salt as chloride (%)	0028	2.1700	0.08000	2.0065	0.10089	0.02938	73	1.62	033.01	0
033	Salt as chloride (%)	0144	2.1800	0.00000	2.0065	0.10089	0.02938	73	1.72	033.03	0
033	Salt as chloride (%)	0539	2.1850	0.01000	2.0065	0.10089	0.02938	73	1.77	033.00	0
033	Salt as chloride (%)	0675	2.2000	0.06000	2.0065	0.10089	0.02938	73	1.92	033.00	0
033	Salt as chloride (%)	0265	2.2550	0.01000	2.0065	0.10089	0.02938	73	2.46	033.03	0
033	Salt as chloride (%)	0682	2.3430	0.00000	2.0065	0.10089	0.02938	73	3.34	033.99	0
033	Salt as chloride (%)	0910	1.6700	0.52000	2.0065	0.10089	0.02938	73	-3.34	033.99	1
034	Selenium (mg / kg (ppm))	0045	1.0220	0.29600	2.2307	0.48622	0.13799	21	-2.49	034.04	0
034	Selenium (mg / kg (ppm))	0047	1.3250	0.07000	2.2307	0.48622	0.13799	21	-1.86	034.52	0
034	Selenium (mg / kg (ppm))	0939	1.7500	0.14000	2.2307	0.48622	0.13799	21	-0.99	034.04	0
034	Selenium (mg / kg (ppm))	0164	1.7900	0.18000	2.2307	0.48622	0.13799	21	-0.91	034.04	0
034	Selenium (mg / kg (ppm))	0910	1.8650	0.01000	2.2307	0.48622	0.13799	21	-0.75	034.52	0
034	Selenium (mg / kg (ppm))	0171	2.0145	0.04300	2.2307	0.48622	0.13799	21	-0.44	034.04	0
034	Selenium (mg / kg (ppm))	0964	2.0705	0.05820	2.2307	0.48622	0.13799	21	-0.33	034.43	0
034	Selenium (mg / kg (ppm))	0553	2.0800	0.04000	2.2307	0.48622	0.13799	21	-0.31	034.53	0
034	Selenium (mg / kg (ppm))	0169	2.1350	0.01000	2.2307	0.48622	0.13799	21	-0.20	034.04	0
034	Selenium (mg / kg (ppm))	0560	2.2095	0.08100	2.2307	0.48622	0.13799	21	-0.04	034.53	0
034	Selenium (mg / kg (ppm))	2053	2.2200	0.04000	2.2307	0.48622	0.13799	21	-0.02	034.04	0
034	Selenium (mg / kg (ppm))	0098	2.2675	0.09700	2.2307	0.48622	0.13799	21	0.08	034.53	0
034	Selenium (mg / kg (ppm))	0208	2.2950	0.09000	2.2307	0.48622	0.13799	21	0.13	034.52	0
034	Selenium (mg / kg (ppm))	0027	2.3130	0.03200	2.2307	0.48622	0.13799	21	0.17	034.53	0
034	Selenium (mg / kg (ppm))	0186	2.3300	0.16000	2.2307	0.48622	0.13799	21	0.20	034.52	0
034	Selenium (mg / kg (ppm))	0190	2.4100	0.04000	2.2307	0.48622	0.13799	21	0.37	034.99	0
034	Selenium (mg / kg (ppm))	0918	2.4672	0.22050	2.2307	0.48622	0.13799	21	0.49	034.53	0
034	Selenium (mg / kg (ppm))	0148	2.8050	0.53000	2.2307	0.48622	0.13799	21	1.18	034.41	0
034	Selenium (mg / kg (ppm))	0036	2.9000	0.20000	2.2307	0.48622	0.13799	21	1.38	034.43	0
034	Selenium (mg / kg (ppm))	0032	3.1600	0.06000	2.2307	0.48622	0.13799	21	1.91	034.53	0
034	Selenium (mg / kg (ppm))	0692	3.2500	0.50000	2.2307	0.48622	0.13799	21	2.10	034.42	0
034	Selenium (mg / kg (ppm))	0160	< 10		2.2307	0.48622	0.13799	21		034.99	5
035	Sodium (%)	2199	0.41500	0.01000	0.73287	0.04682	0.01959	107	-6.79	035.05	0
035	Sodium (%)	0536	0.62500	0.01000	0.73287	0.04682	0.01959	107	-2.30	035.31	0
035	Sodium (%)	0609	0.64000	0.02000	0.73287	0.04682	0.01959	107	-1.98	035.32	0
035	Sodium (%)	2192	0.64000	0.04000	0.73287	0.04682	0.01959	107	-1.98	035.43	0
035	Sodium (%)	0692	0.65500	0.11000	0.73287	0.04682	0.01959	107	-1.66	035.42	0
035	Sodium (%)	0298	0.66000	0.02000	0.73287	0.04682	0.01959	107	-1.56	035.41	0
035	Sodium (%)	0042	0.66200	0.01800	0.73287	0.04682	0.01959	107	-1.51	035.42	0
035	Sodium (%)	0650	0.66300	0.02000	0.73287	0.04682	0.01959	107	-1.49	035.31	0
035	Sodium (%)	0175	0.66500	0.01000	0.73287	0.04682	0.01959	107	-1.45	035.31	0
035	Sodium (%)	0049	0.66500	0.03000	0.73287	0.04682	0.01959	107	-1.45	035.41	0
035	Sodium (%)	0035	0.66500	0.01000	0.73287	0.04682	0.01959	107	-1.45	035.43	0
035	Sodium (%)	0345	0.66750	0.00500	0.73287	0.04682	0.01959	107	-1.40	035.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
035	Sodium (%)	0043	0.66900	0.01200	0.73287	0.04682	0.01959	107	-1.36	035.43	0
035	Sodium (%)	0868	0.67400	0.00200	0.73287	0.04682	0.01959	107	-1.26	035.01	0
035	Sodium (%)	0169	0.67500	0.03000	0.73287	0.04682	0.01959	107	-1.24	035.32	0
035	Sodium (%)	0144	0.67500	0.03000	0.73287	0.04682	0.01959	107	-1.24	035.41	0
035	Sodium (%)	0035	0.67500	0.01000	0.73287	0.04682	0.01959	107	-1.24	035.42	0
035	Sodium (%)	0142	0.67685	0.00450	0.73287	0.04682	0.01959	107	-1.20	035.31	0
035	Sodium (%)	2146	0.67700	0.00400	0.73287	0.04682	0.01959	107	-1.19	035.31	0
035	Sodium (%)	2022	0.67800	0.00600	0.73287	0.04682	0.01959	107	-1.17	035.31	0
035	Sodium (%)	0353	0.68500	0.01000	0.73287	0.04682	0.01959	107	-1.02	035.43	0
035	Sodium (%)	0045	0.68600	0.04200	0.73287	0.04682	0.01959	107	-1.00	035.42	0
035	Sodium (%)	0300	0.69000	0.04000	0.73287	0.04682	0.01959	107	-0.92	035.33	0
035	Sodium (%)	0098	0.69000	0.04000	0.73287	0.04682	0.01959	107	-0.92	035.43	0
035	Sodium (%)	0910	0.70000	0.02000	0.73287	0.04682	0.01959	107	-0.70	035.41	0
035	Sodium (%)	0265	0.70000	0.10000	0.73287	0.04682	0.01959	107	-0.70	035.42	0
035	Sodium (%)	0171	0.70200	0.02800	0.73287	0.04682	0.01959	107	-0.66	035.41	0
035	Sodium (%)	0629	0.70500	0.01000	0.73287	0.04682	0.01959	107	-0.60	035.31	0
035	Sodium (%)	0629	0.70500	0.03000	0.73287	0.04682	0.01959	107	-0.60	035.43	0
035	Sodium (%)	0186	0.70800	0.01800	0.73287	0.04682	0.01959	107	-0.53	035.42	0
035	Sodium (%)	0019	0.71000	0.00000	0.73287	0.04682	0.01959	107	-0.49	035.41	0
035	Sodium (%)	0278	0.71000	0.02000	0.73287	0.04682	0.01959	107	-0.49	035.42	0
035	Sodium (%)	0884	0.71500	0.07000	0.73287	0.04682	0.01959	107	-0.38	035.02	0
035	Sodium (%)	0148	0.71520	0.00560	0.73287	0.04682	0.01959	107	-0.38	035.41	0
035	Sodium (%)	0889	0.71940	0.07280	0.73287	0.04682	0.01959	107	-0.29	035.99	0
035	Sodium (%)	0683	0.72000	0.02000	0.73287	0.04682	0.01959	107	-0.27	035.31	0
035	Sodium (%)	0511	0.72000	0.00000	0.73287	0.04682	0.01959	107	-0.27	035.41	0
035	Sodium (%)	0186	0.72100	0.00800	0.73287	0.04682	0.01959	107	-0.25	035.52	0
035	Sodium (%)	0504	0.72150	0.01900	0.73287	0.04682	0.01959	107	-0.24	035.42	0
035	Sodium (%)	0563	0.72200	0.01740	0.73287	0.04682	0.01959	107	-0.23	035.31	0
035	Sodium (%)	0653	0.72650	0.00500	0.73287	0.04682	0.01959	107	-0.14	035.41	0
035	Sodium (%)	0202	0.72750	0.07500	0.73287	0.04682	0.01959	107	-0.11	035.43	0
035	Sodium (%)	0032	0.72905	0.00930	0.73287	0.04682	0.01959	107	-0.08	035.43	0
035	Sodium (%)	2196	0.73000	0.00000	0.73287	0.04682	0.01959	107	-0.06	035.31	0
035	Sodium (%)	0098	0.73000	0.04000	0.73287	0.04682	0.01959	107	-0.06	035.41	0
035	Sodium (%)	0205	0.73000	0.00000	0.73287	0.04682	0.01959	107	-0.06	035.42	0
035	Sodium (%)	0263	0.73275	0.00250	0.73287	0.04682	0.01959	107	0.00	035.42	0
035	Sodium (%)	0598	0.73435	0.00710	0.73287	0.04682	0.01959	107	0.03	035.41	0
035	Sodium (%)	0668	0.73450	0.00900	0.73287	0.04682	0.01959	107	0.03	035.43	0
035	Sodium (%)	0229	0.73500	0.03000	0.73287	0.04682	0.01959	107	0.05	035.41	0
035	Sodium (%)	0675	0.73500	0.03000	0.73287	0.04682	0.01959	107	0.05	035.43	0
035	Sodium (%)	0870	0.73680	0.02020	0.73287	0.04682	0.01959	107	0.08	035.42	0
035	Sodium (%)	0407	0.73770	0.02140	0.73287	0.04682	0.01959	107	0.10	035.41	0
035	Sodium (%)	0939	0.74000	0.02000	0.73287	0.04682	0.01959	107	0.15	035.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
035	Sodium (%)	0878	0.74000	0.04000	0.73287	0.04682	0.01959	107	0.15	035.41	0
035	Sodium (%)	0861	0.74000	0.06800	0.73287	0.04682	0.01959	107	0.15	035.43	0
035	Sodium (%)	0208	0.74150	0.01700	0.73287	0.04682	0.01959	107	0.18	035.41	0
035	Sodium (%)	0505	0.74200	0.00000	0.73287	0.04682	0.01959	107	0.19	035.41	0
035	Sodium (%)	0948	0.74450	0.00500	0.73287	0.04682	0.01959	107	0.25	035.31	0
035	Sodium (%)	0187	0.74485	0.00170	0.73287	0.04682	0.01959	107	0.26	035.42	0
035	Sodium (%)	0242	0.74500	0.05000	0.73287	0.04682	0.01959	107	0.26	035.99	0
035	Sodium (%)	0017	0.74700	0.00200	0.73287	0.04682	0.01959	107	0.30	035.43	0
035	Sodium (%)	0164	0.74750	0.01300	0.73287	0.04682	0.01959	107	0.31	035.41	0
035	Sodium (%)	0874	0.74850	0.04100	0.73287	0.04682	0.01959	107	0.33	035.31	0
035	Sodium (%)	2006	0.74900	0.00200	0.73287	0.04682	0.01959	107	0.34	035.01	0
035	Sodium (%)	0682	0.75000	0.00000	0.73287	0.04682	0.01959	107	0.37	035.43	0
035	Sodium (%)	0006	0.75100	0.00600	0.73287	0.04682	0.01959	107	0.39	035.43	0
035	Sodium (%)	0520	0.75140	0.01160	0.73287	0.04682	0.01959	107	0.40	035.41	0
035	Sodium (%)	0881	0.75450	0.01900	0.73287	0.04682	0.01959	107	0.46	035.99	0
035	Sodium (%)	0504	0.75500	0.01000	0.73287	0.04682	0.01959	107	0.47	035.05	0
035	Sodium (%)	0354	0.75500	0.01000	0.73287	0.04682	0.01959	107	0.47	035.31	0
035	Sodium (%)	0868	0.75500	0.01000	0.73287	0.04682	0.01959	107	0.47	035.31	0
035	Sodium (%)	0560	0.75520	0.09140	0.73287	0.04682	0.01959	107	0.48	035.42	0
035	Sodium (%)	0208	0.75550	0.01100	0.73287	0.04682	0.01959	107	0.48	035.31	0
035	Sodium (%)	0358	0.76000	0.00000	0.73287	0.04682	0.01959	107	0.58	035.41	0
035	Sodium (%)	0413	0.76000	0.00000	0.73287	0.04682	0.01959	107	0.58	035.41	0
035	Sodium (%)	0297	0.76000	0.02000	0.73287	0.04682	0.01959	107	0.58	035.43	0
035	Sodium (%)	0100	0.76000	0.02000	0.73287	0.04682	0.01959	107	0.58	035.99	0
035	Sodium (%)	0693	0.76050	0.00100	0.73287	0.04682	0.01959	107	0.59	035.42	0
035	Sodium (%)	0510	0.76150	0.00300	0.73287	0.04682	0.01959	107	0.61	035.43	0
035	Sodium (%)	0139	0.76350	0.01300	0.73287	0.04682	0.01959	107	0.65	035.41	0
035	Sodium (%)	0572	0.76400	0.03200	0.73287	0.04682	0.01959	107	0.66	035.53	0
035	Sodium (%)	0037	0.76500	0.01000	0.73287	0.04682	0.01959	107	0.69	035.05	0
035	Sodium (%)	0294	0.76500	0.01000	0.73287	0.04682	0.01959	107	0.69	035.42	0
035	Sodium (%)	0553	0.76650	0.00500	0.73287	0.04682	0.01959	107	0.72	035.53	0
035	Sodium (%)	0918	0.76900	0.05800	0.73287	0.04682	0.01959	107	0.77	035.43	0
035	Sodium (%)	0038	0.76915	0.08230	0.73287	0.04682	0.01959	107	0.77	035.31	0
035	Sodium (%)	0089	0.77000	0.00000	0.73287	0.04682	0.01959	107	0.79	035.43	0
035	Sodium (%)	0590	0.77000	0.02000	0.73287	0.04682	0.01959	107	0.79	035.99	0
035	Sodium (%)	0940	0.77500	0.01000	0.73287	0.04682	0.01959	107	0.90	035.99	0
035	Sodium (%)	0226	0.77800	0.01200	0.73287	0.04682	0.01959	107	0.96	035.41	0
035	Sodium (%)	0154	0.77955	0.03870	0.73287	0.04682	0.01959	107	1.00	035.52	0
035	Sodium (%)	0689	0.78000	0.00000	0.73287	0.04682	0.01959	107	1.01	035.31	0
035	Sodium (%)	0720	0.78000	0.02000	0.73287	0.04682	0.01959	107	1.01	035.41	0
035	Sodium (%)	0610	0.78000	0.00000	0.73287	0.04682	0.01959	107	1.01	035.43	0
035	Sodium (%)	2089	0.78000	0.00000	0.73287	0.04682	0.01959	107	1.01	035.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
035	Sodium (%)	0108	0.78500	0.01000	0.73287	0.04682	0.01959	107	1.11	035.05	0
035	Sodium (%)	0038	0.79200	0.00800	0.73287	0.04682	0.01959	107	1.26	035.43	0
035	Sodium (%)	2188	0.79295	0.00070	0.73287	0.04682	0.01959	107	1.28	035.31	0
035	Sodium (%)	0964	0.79490	0.01140	0.73287	0.04682	0.01959	107	1.32	035.43	0
035	Sodium (%)	2126	0.79500	0.01000	0.73287	0.04682	0.01959	107	1.33	035.05	0
035	Sodium (%)	0405	0.79500	0.01000	0.73287	0.04682	0.01959	107	1.33	035.41	0
035	Sodium (%)	0726	0.79620	0.00100	0.73287	0.04682	0.01959	107	1.35	035.42	0
035	Sodium (%)	0190	0.80000	0.02000	0.73287	0.04682	0.01959	107	1.43	035.42	0
035	Sodium (%)	0083	0.80500	0.01000	0.73287	0.04682	0.01959	107	1.54	035.43	0
035	Sodium (%)	2161	0.81500	0.01000	0.73287	0.04682	0.01959	107	1.75	035.99	0
035	Sodium (%)	0425	0.83000	0.00000	0.73287	0.04682	0.01959	107	2.07	035.43	0
035	Sodium (%)	2195	7,213.0	140.00	0.73287	0.04682	0.01959	107	154034.70	035.31	2
036	Sulfur (%)	2161	0.14500	0.01000	0.29543	0.02829	0.01015	44	-5.32	036.00	0
036	Sulfur (%)	2053	0.22500	0.01000	0.29543	0.02829	0.01015	44	-2.49	036.04	0
036	Sulfur (%)	0035	0.23500	0.01000	0.29543	0.02829	0.01015	44	-2.14	036.43	0
036	Sulfur (%)	0870	0.24955	0.00630	0.29543	0.02829	0.01015	44	-1.62	036.42	0
036	Sulfur (%)	0045	0.25350	0.02700	0.29543	0.02829	0.01015	44	-1.48	036.42	0
036	Sulfur (%)	0353	0.26000	0.02000	0.29543	0.02829	0.01015	44	-1.25	036.43	0
036	Sulfur (%)	0035	0.26500	0.01000	0.29543	0.02829	0.01015	44	-1.08	036.42	0
036	Sulfur (%)	0910	0.27170	0.00160	0.29543	0.02829	0.01015	44	-0.84	036.52	0
036	Sulfur (%)	0171	0.27300	0.00000	0.29543	0.02829	0.01015	44	-0.79	036.42	0
036	Sulfur (%)	0345	0.27500	0.00600	0.29543	0.02829	0.01015	44	-0.72	036.43	0
036	Sulfur (%)	2089	0.27500	0.01000	0.29543	0.02829	0.01015	44	-0.72	036.43	0
036	Sulfur (%)	0265	0.28000	0.00000	0.29543	0.02829	0.01015	44	-0.55	036.42	0
036	Sulfur (%)	0889	0.28140	0.00360	0.29543	0.02829	0.01015	44	-0.50	036.99	0
036	Sulfur (%)	0407	0.28820	0.00320	0.29543	0.02829	0.01015	44	-0.26	036.42	0
036	Sulfur (%)	0668	0.28850	0.01300	0.29543	0.02829	0.01015	44	-0.24	036.43	0
036	Sulfur (%)	0693	0.29000	0.00000	0.29543	0.02829	0.01015	44	-0.19	036.42	0
036	Sulfur (%)	0164	0.29000	0.02000	0.29543	0.02829	0.01015	44	-0.19	036.42	0
036	Sulfur (%)	0357	0.29105	0.00030	0.29543	0.02829	0.01015	44	-0.15	036.42	0
036	Sulfur (%)	0186	0.29300	0.00400	0.29543	0.02829	0.01015	44	-0.09	036.52	0
036	Sulfur (%)	0202	0.29350	0.00100	0.29543	0.02829	0.01015	44	-0.07	036.43	0
036	Sulfur (%)	0229	0.29500	0.01000	0.29543	0.02829	0.01015	44	-0.02	036.42	0
036	Sulfur (%)	0297	0.29500	0.01000	0.29543	0.02829	0.01015	44	-0.02	036.43	0
036	Sulfur (%)	0692	0.29500	0.03000	0.29543	0.02829	0.01015	44	-0.02	036.42	0
036	Sulfur (%)	0187	0.29690	0.00180	0.29543	0.02829	0.01015	44	0.05	036.42	0
036	Sulfur (%)	0229	0.30000	0.02000	0.29543	0.02829	0.01015	44	0.16	036.04	0
036	Sulfur (%)	0033	0.30000	0.02000	0.29543	0.02829	0.01015	44	0.16	036.43	0
036	Sulfur (%)	0629	0.30000	0.00000	0.29543	0.02829	0.01015	44	0.16	036.43	0
036	Sulfur (%)	0726	0.30260	0.00040	0.29543	0.02829	0.01015	44	0.25	036.42	0
036	Sulfur (%)	0226	0.30500	0.03000	0.29543	0.02829	0.01015	44	0.34	036.04	0
036	Sulfur (%)	0918	0.30900	0.01000	0.29543	0.02829	0.01015	44	0.48	036.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
036	Sulfur (%)	0098	0.30920	0.00400	0.29543	0.02829	0.01015	44	0.49	036.43	0
036	Sulfur (%)	0190	0.31000	0.02000	0.29543	0.02829	0.01015	44	0.52	036.42	0
036	Sulfur (%)	0169	0.31000	0.00000	0.29543	0.02829	0.01015	44	0.52	036.43	0
036	Sulfur (%)	0242	0.31000	0.00000	0.29543	0.02829	0.01015	44	0.52	036.99	0
036	Sulfur (%)	0560	0.31220	0.03920	0.29543	0.02829	0.01015	44	0.59	036.42	0
036	Sulfur (%)	0186	0.31250	0.00100	0.29543	0.02829	0.01015	44	0.60	036.42	0
036	Sulfur (%)	0708	0.31600	0.00600	0.29543	0.02829	0.01015	44	0.73	036.42	0
036	Sulfur (%)	0964	0.31665	0.00210	0.29543	0.02829	0.01015	44	0.75	036.43	0
036	Sulfur (%)	0037	0.33500	0.01000	0.29543	0.02829	0.01015	44	1.40	036.42	0
036	Sulfur (%)	0038	0.33650	0.01300	0.29543	0.02829	0.01015	44	1.45	036.43	0
036	Sulfur (%)	0278	0.34000	0.02000	0.29543	0.02829	0.01015	44	1.58	036.42	0
036	Sulfur (%)	0032	0.35380	0.03300	0.29543	0.02829	0.01015	44	2.06	036.43	0
036	Sulfur (%)	0510	0.37000	0.00000	0.29543	0.02829	0.01015	44	2.64	036.43	0
036	Sulfur (%)	0294	0.42500	0.01000	0.29543	0.02829	0.01015	44	4.58	036.42	0
036	Sulfur (%)	0553	0.31800	0.07200	0.29543	0.02829	0.01015	44	0.80	036.53	1
037	Zinc (mg / kg (ppm))	0405	383.00	0.20000	634.52	42.354	16.967	98	-5.94	037.41	0
037	Zinc (mg / kg (ppm))	0609	429.00	0.00000	634.52	42.354	16.967	98	-4.85	037.32	0
037	Zinc (mg / kg (ppm))	0596	436.15	13.300	634.52	42.354	16.967	98	-4.68	037.31	0
037	Zinc (mg / kg (ppm))	0889	441.20	0.91060	634.52	42.354	16.967	98	-4.56	037.99	0
037	Zinc (mg / kg (ppm))	0045	545.00	64.000	634.52	42.354	16.967	98	-2.11	037.42	0
037	Zinc (mg / kg (ppm))	0049	545.73	36.860	634.52	42.354	16.967	98	-2.10	037.41	0
037	Zinc (mg / kg (ppm))	2192	549.29	26.130	634.52	42.354	16.967	98	-2.01	037.43	0
037	Zinc (mg / kg (ppm))	0345	556.00	8.0000	634.52	42.354	16.967	98	-1.85	037.43	0
037	Zinc (mg / kg (ppm))	0353	563.00	14.000	634.52	42.354	16.967	98	-1.69	037.43	0
037	Zinc (mg / kg (ppm))	0668	566.50	7.0000	634.52	42.354	16.967	98	-1.61	037.43	0
037	Zinc (mg / kg (ppm))	2129	581.80	7.8000	634.52	42.354	16.967	98	-1.24	037.99	0
037	Zinc (mg / kg (ppm))	0034	582.00	8.4000	634.52	42.354	16.967	98	-1.24	037.43	0
037	Zinc (mg / kg (ppm))	0878	583.50	7.0000	634.52	42.354	16.967	98	-1.20	037.41	0
037	Zinc (mg / kg (ppm))	0720	587.45	6.7500	634.52	42.354	16.967	98	-1.11	037.41	0
037	Zinc (mg / kg (ppm))	0278	590.30	4.0000	634.52	42.354	16.967	98	-1.04	037.42	0
037	Zinc (mg / kg (ppm))	0884	590.80	4.8000	634.52	42.354	16.967	98	-1.03	037.31	0
037	Zinc (mg / kg (ppm))	2022	592.00	0.00000	634.52	42.354	16.967	98	-1.00	037.31	0
037	Zinc (mg / kg (ppm))	0874	596.00	22.000	634.52	42.354	16.967	98	-0.91	037.31	0
037	Zinc (mg / kg (ppm))	0202	596.66	8.1600	634.52	42.354	16.967	98	-0.89	037.43	0
037	Zinc (mg / kg (ppm))	0511	598.00	8.0000	634.52	42.354	16.967	98	-0.86	037.41	0
037	Zinc (mg / kg (ppm))	0520	602.50	59.000	634.52	42.354	16.967	98	-0.76	037.41	0
037	Zinc (mg / kg (ppm))	0038	603.10	20.200	634.52	42.354	16.967	98	-0.74	037.32	0
037	Zinc (mg / kg (ppm))	0164	605.50	9.0000	634.52	42.354	16.967	98	-0.69	037.41	0
037	Zinc (mg / kg (ppm))	0043	608.00	34.000	634.52	42.354	16.967	98	-0.63	037.43	0
037	Zinc (mg / kg (ppm))	0033	611.50	21.000	634.52	42.354	16.967	98	-0.54	037.43	0
037	Zinc (mg / kg (ppm))	0187	612.78	0.85000	634.52	42.354	16.967	98	-0.51	037.42	0
037	Zinc (mg / kg (ppm))	2188	613.73	6.1485	634.52	42.354	16.967	98	-0.49	037.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
037	Zinc (mg / kg (ppm))	0629	615.00	4.0000	634.52	42.354	16.967	98	-0.46	037.31	0
037	Zinc (mg / kg (ppm))	0171	616.10	3.2000	634.52	42.354	16.967	98	-0.43	037.41	0
037	Zinc (mg / kg (ppm))	0186	617.50	1.0000	634.52	42.354	16.967	98	-0.40	037.52	0
037	Zinc (mg / kg (ppm))	2128	618.18	59.492	634.52	42.354	16.967	98	-0.39	037.32	0
037	Zinc (mg / kg (ppm))	0505	618.55	18.900	634.52	42.354	16.967	98	-0.38	037.41	0
037	Zinc (mg / kg (ppm))	2144	618.84	20.260	634.52	42.354	16.967	98	-0.37	037.99	0
037	Zinc (mg / kg (ppm))	0653	619.25	0.30000	634.52	42.354	16.967	98	-0.36	037.41	0
037	Zinc (mg / kg (ppm))	0042	619.50	7.0000	634.52	42.354	16.967	98	-0.35	037.42	0
037	Zinc (mg / kg (ppm))	0939	622.14	3.1700	634.52	42.354	16.967	98	-0.29	037.31	0
037	Zinc (mg / kg (ppm))	0148	622.30	6.2000	634.52	42.354	16.967	98	-0.29	037.41	0
037	Zinc (mg / kg (ppm))	0629	624.50	13.000	634.52	42.354	16.967	98	-0.24	037.43	0
037	Zinc (mg / kg (ppm))	2089	625.41	7.3600	634.52	42.354	16.967	98	-0.22	037.43	0
037	Zinc (mg / kg (ppm))	0098	626.40	10.800	634.52	42.354	16.967	98	-0.19	037.41	0
037	Zinc (mg / kg (ppm))	0968	627.00	8.0000	634.52	42.354	16.967	98	-0.18	037.43	0
037	Zinc (mg / kg (ppm))	0358	627.22	1.5500	634.52	42.354	16.967	98	-0.17	037.41	0
037	Zinc (mg / kg (ppm))	0017	628.15	28.862	634.52	42.354	16.967	98	-0.15	037.43	0
037	Zinc (mg / kg (ppm))	0003	629.50	13.000	634.52	42.354	16.967	98	-0.12	037.41	0
037	Zinc (mg / kg (ppm))	0354	629.69	4.7100	634.52	42.354	16.967	98	-0.11	037.31	0
037	Zinc (mg / kg (ppm))	2195	630.50	19.000	634.52	42.354	16.967	98	-0.09	037.31	0
037	Zinc (mg / kg (ppm))	0242	631.00	26.000	634.52	42.354	16.967	98	-0.08	037.99	0
037	Zinc (mg / kg (ppm))	0910	631.50	43.000	634.52	42.354	16.967	98	-0.07	037.41	0
037	Zinc (mg / kg (ppm))	0610	632.00	0.00000	634.52	42.354	16.967	98	-0.06	037.43	0
037	Zinc (mg / kg (ppm))	0504	632.40	35.600	634.52	42.354	16.967	98	-0.05	037.32	0
037	Zinc (mg / kg (ppm))	0590	633.50	25.000	634.52	42.354	16.967	98	-0.02	037.99	0
037	Zinc (mg / kg (ppm))	0689	636.40	27.400	634.52	42.354	16.967	98	0.04	037.31	0
037	Zinc (mg / kg (ppm))	0186	636.50	3.0000	634.52	42.354	16.967	98	0.05	037.42	0
037	Zinc (mg / kg (ppm))	2196	637.20	0.00000	634.52	42.354	16.967	98	0.06	037.31	0
037	Zinc (mg / kg (ppm))	0357	637.94	0.51000	634.52	42.354	16.967	98	0.08	037.42	0
037	Zinc (mg / kg (ppm))	0868	639.33	0.21000	634.52	42.354	16.967	98	0.11	037.31	0
037	Zinc (mg / kg (ppm))	0294	640.09	15.420	634.52	42.354	16.967	98	0.13	037.42	0
037	Zinc (mg / kg (ppm))	0560	641.00	60.000	634.52	42.354	16.967	98	0.15	037.42	0
037	Zinc (mg / kg (ppm))	0598	644.90	31.000	634.52	42.354	16.967	98	0.25	037.41	0
037	Zinc (mg / kg (ppm))	0175	645.00	10.000	634.52	42.354	16.967	98	0.25	037.31	0
037	Zinc (mg / kg (ppm))	0510	647.00	12.000	634.52	42.354	16.967	98	0.29	037.43	0
037	Zinc (mg / kg (ppm))	0265	648.00	60.000	634.52	42.354	16.967	98	0.32	037.42	0
037	Zinc (mg / kg (ppm))	0964	648.86	39.601	634.52	42.354	16.967	98	0.34	037.43	0
037	Zinc (mg / kg (ppm))	2145	652.75	9.9000	634.52	42.354	16.967	98	0.43	037.41	0
037	Zinc (mg / kg (ppm))	0074	653.00	24.000	634.52	42.354	16.967	98	0.44	037.41	0
037	Zinc (mg / kg (ppm))	0083	654.00	14.000	634.52	42.354	16.967	98	0.46	037.43	0
037	Zinc (mg / kg (ppm))	0019	654.05	28.300	634.52	42.354	16.967	98	0.46	037.41	0
037	Zinc (mg / kg (ppm))	0001	655.25	19.700	634.52	42.354	16.967	98	0.49	037.31	0
037	Zinc (mg / kg (ppm))	0563	658.22	38.419	634.52	42.354	16.967	98	0.56	037.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
037	Zinc (mg / kg (ppm))	0572	658.50	7.0000	634.52	42.354	16.967	98	0.57	037.53	0
037	Zinc (mg / kg (ppm))	0208	660.00	16.000	634.52	42.354	16.967	98	0.60	037.31	0
037	Zinc (mg / kg (ppm))	0861	661.00	44.000	634.52	42.354	16.967	98	0.63	037.43	0
037	Zinc (mg / kg (ppm))	0407	663.38	3.7524	634.52	42.354	16.967	98	0.68	037.41	0
037	Zinc (mg / kg (ppm))	0032	663.55	26.500	634.52	42.354	16.967	98	0.69	037.43	0
037	Zinc (mg / kg (ppm))	0870	664.65	13.700	634.52	42.354	16.967	98	0.71	037.42	0
037	Zinc (mg / kg (ppm))	0098	664.65	51.300	634.52	42.354	16.967	98	0.71	037.43	0
037	Zinc (mg / kg (ppm))	0010	665.00	10.000	634.52	42.354	16.967	98	0.72	037.33	0
037	Zinc (mg / kg (ppm))	0413	665.00	16.000	634.52	42.354	16.967	98	0.72	037.42	0
037	Zinc (mg / kg (ppm))	0229	668.00	3.6000	634.52	42.354	16.967	98	0.79	037.41	0
037	Zinc (mg / kg (ppm))	0037	670.20	36.200	634.52	42.354	16.967	98	0.84	037.42	0
037	Zinc (mg / kg (ppm))	0675	672.35	33.840	634.52	42.354	16.967	98	0.89	037.43	0
037	Zinc (mg / kg (ppm))	0425	674.20	1.2000	634.52	42.354	16.967	98	0.94	037.43	0
037	Zinc (mg / kg (ppm))	0682	675.42	0.00000	634.52	42.354	16.967	98	0.97	037.43	0
037	Zinc (mg / kg (ppm))	0226	675.50	25.000	634.52	42.354	16.967	98	0.97	037.41	0
037	Zinc (mg / kg (ppm))	0190	678.56	2.6800	634.52	42.354	16.967	98	1.04	037.42	0
037	Zinc (mg / kg (ppm))	0027	679.01	13.601	634.52	42.354	16.967	98	1.05	037.43	0
037	Zinc (mg / kg (ppm))	0693	683.50	11.000	634.52	42.354	16.967	98	1.16	037.42	0
037	Zinc (mg / kg (ppm))	0208	684.00	13.600	634.52	42.354	16.967	98	1.17	037.41	0
037	Zinc (mg / kg (ppm))	0297	690.50	25.000	634.52	42.354	16.967	98	1.32	037.43	0
037	Zinc (mg / kg (ppm))	0553	693.50	29.000	634.52	42.354	16.967	98	1.39	037.53	0
037	Zinc (mg / kg (ppm))	0726	695.65	2.3000	634.52	42.354	16.967	98	1.44	037.42	0
037	Zinc (mg / kg (ppm))	2053	697.50	43.000	634.52	42.354	16.967	98	1.49	037.42	0
037	Zinc (mg / kg (ppm))	2161	709.50	1.0000	634.52	42.354	16.967	98	1.77	037.99	0
037	Zinc (mg / kg (ppm))	0100	715.00	0.00000	634.52	42.354	16.967	98	1.90	037.99	0
037	Zinc (mg / kg (ppm))	0918	724.57	25.170	634.52	42.354	16.967	98	2.13	037.43	0
037	Zinc (mg / kg (ppm))	0948	729.69	0.01000	634.52	42.354	16.967	98	2.25	037.33	0
037	Zinc (mg / kg (ppm))	0263	805.13	45.980	634.52	42.354	16.967	98	4.03	037.42	0
037	Zinc (mg / kg (ppm))	0169	827.00	12.000	634.52	42.354	16.967	98	4.54	037.43	0
037	Zinc (mg / kg (ppm))	0047	272.15	260.50	634.52	42.354	16.967	98	-8.56	037.52	1
037	Zinc (mg / kg (ppm))	0692	570.00	148.00	634.52	42.354	16.967	98	-1.52	037.42	1
037	Zinc (mg / kg (ppm))	2178	781.50	135.00	634.52	42.354	16.967	98	3.47	037.33	1
037	Zinc (mg / kg (ppm))	0536	80.625	3.7500	634.52	42.354	16.967	98	-13.08	037.31	2
038	Molybdenum (mg / kg (ppm))	0353	1.4275	0.03470	2.0476	0.42915	0.07626	19	-1.44	038.43	0
038	Molybdenum (mg / kg (ppm))	0171	1.4805	0.01900	2.0476	0.42915	0.07626	19	-1.32	038.41	0
038	Molybdenum (mg / kg (ppm))	0345	1.6750	0.15000	2.0476	0.42915	0.07626	19	-0.87	038.43	0
038	Molybdenum (mg / kg (ppm))	0910	1.7550	0.21000	2.0476	0.42915	0.07626	19	-0.68	038.52	0
038	Molybdenum (mg / kg (ppm))	0169	1.7650	0.01000	2.0476	0.42915	0.07626	19	-0.66	038.43	0
038	Molybdenum (mg / kg (ppm))	0098	1.7905	0.23100	2.0476	0.42915	0.07626	19	-0.60	038.53	0
038	Molybdenum (mg / kg (ppm))	0278	1.8300	0.02000	2.0476	0.42915	0.07626	19	-0.51	038.42	0
038	Molybdenum (mg / kg (ppm))	0226	1.8350	0.01000	2.0476	0.42915	0.07626	19	-0.50	038.41	0
038	Molybdenum (mg / kg (ppm))	0148	1.9700	0.02000	2.0476	0.42915	0.07626	19	-0.18	038.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
038	Molybdenum (mg / kg (ppm))	0297	1.9800	0.02000	2.0476	0.42915	0.07626	19	-0.16	038.43	0
038	Molybdenum (mg / kg (ppm))	0693	2.0000	0.00000	2.0476	0.42915	0.07626	19	-0.11	038.42	0
038	Molybdenum (mg / kg (ppm))	0186	2.1250	0.03000	2.0476	0.42915	0.07626	19	0.18	038.52	0
038	Molybdenum (mg / kg (ppm))	0918	2.2620	0.16620	2.0476	0.42915	0.07626	19	0.50	038.53	0
038	Molybdenum (mg / kg (ppm))	0553	2.2650	0.03000	2.0476	0.42915	0.07626	19	0.51	038.53	0
038	Molybdenum (mg / kg (ppm))	0038	2.4350	0.07000	2.0476	0.42915	0.07626	19	0.90	038.43	0
038	Molybdenum (mg / kg (ppm))	0045	2.4750	0.39000	2.0476	0.42915	0.07626	19	1.00	038.42	0
038	Molybdenum (mg / kg (ppm))	0032	2.5420	0.03800	2.0476	0.42915	0.07626	19	1.15	038.43	0
038	Molybdenum (mg / kg (ppm))	0510	2.6000	0.00000	2.0476	0.42915	0.07626	19	1.29	038.43	0
038	Molybdenum (mg / kg (ppm))	0265	3.0000	0.00000	2.0476	0.42915	0.07626	19	2.22	038.42	0
038	Molybdenum (mg / kg (ppm))	0964	2.2090	0.49130	2.0476	0.42915	0.07626	19	0.38	038.43	1
038	Molybdenum (mg / kg (ppm))	2161	< 100		2.0476	0.42915	0.07626	19		038.99	5
040	Barium (mg / kg (ppm))	0918	4.2307	0.19890				2		040.53	0
040	Barium (mg / kg (ppm))	0560	4.9400	0.24000				2		040.42	0
041	Vanadium (mg / kg (ppm))	0553	4.1500	0.26000				1		041.53	0
042	Chloride (%)	0297	1.1900	0.04000	1.2448	0.07153	0.03500	4		042.00	0
042	Chloride (%)	0948	1.2190	0.00000	1.2448	0.07153	0.03500	4		042.00	0
042	Chloride (%)	0265	1.2200	0.00000	1.2448	0.07153	0.03500	4		042.01	0
042	Chloride (%)	2199	1.3500	0.10000	1.2448	0.07153	0.03500	4		042.00	0
101	Choline Chloride (mg / kg (ppm))	0675	985.81	149.84	2,062.3	728.59	131.71	4		101.00	0
101	Choline Chloride (mg / kg (ppm))	0939	2,243.5	237.00	2,062.3	728.59	131.71	4		101.01	0
101	Choline Chloride (mg / kg (ppm))	0227	2,510.0	80.000	2,062.3	728.59	131.71	4		101.00	0
101	Choline Chloride (mg / kg (ppm))	0160	2,510.0	60.000	2,062.3	728.59	131.71	4		101.99	0
102	Niacin (mg / kg (ppm))	0160	113.00	8.0000				2		102.02	0
102	Niacin (mg / kg (ppm))	0227	187.00	6.0000				2		102.01	0
103	Pantothenic Acid (mg / kg (ppm))	0227	86.100	1.6000				2		103.01	0
103	Pantothenic Acid (mg / kg (ppm))	0160	133.50	5.0000				2		103.02	0
104	Riboflavin (mg / kg (ppm))	0160	22.100	2.0000	30.568	6.0727	1.3461	5		104.03	0
104	Riboflavin (mg / kg (ppm))	0910	26.130	2.1600	30.568	6.0727	1.3461	5		104.03	0
104	Riboflavin (mg / kg (ppm))	0227	34.300	0.20000	30.568	6.0727	1.3461	5		104.00	0
104	Riboflavin (mg / kg (ppm))	0171	34.950	1.5000	30.568	6.0727	1.3461	5		104.00	0
104	Riboflavin (mg / kg (ppm))	0027	35.361	0.87040	30.568	6.0727	1.3461	5		104.03	0
105	Thiamine (mg / kg (ppm))	0629	2.2300	0.02000	6.9238	3.4183	0.78250	4		105.00	0
105	Thiamine (mg / kg (ppm))	0910	6.6800	0.40000	6.9238	3.4183	0.78250	4		105.00	0
105	Thiamine (mg / kg (ppm))	0227	8.7700	1.5400	6.9238	3.4183	0.78250	4		105.01	0
105	Thiamine (mg / kg (ppm))	0160	10.015	1.1700	6.9238	3.4183	0.78250	4		105.00	0
106	Vitamin A (KU / kg)	2161	11.140	0.52000	24.468	7.2701	3.4444	22	-1.83	106.02	0
106	Vitamin A (KU / kg)	0941	15.450	1.3000	24.468	7.2701	3.4444	22	-1.24	106.02	0
106	Vitamin A (KU / kg)	0098	16.300	5.4000	24.468	7.2701	3.4444	22	-1.12	106.01	0
106	Vitamin A (KU / kg)	0038	16.485	1.8300	24.468	7.2701	3.4444	22	-1.10	106.02	0
106	Vitamin A (KU / kg)	2053	17.450	5.9000	24.468	7.2701	3.4444	22	-0.97	106.02	0
106	Vitamin A (KU / kg)	0004	20.295	3.1100	24.468	7.2701	3.4444	22	-0.57	106.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
106	Vitamin A (KU / kg)	0875	21.600	0.80000	24.468	7.2701	3.4444	22	-0.39	106.02	0
106	Vitamin A (KU / kg)	0227	21.850	0.70000	24.468	7.2701	3.4444	22	-0.36	106.02	0
106	Vitamin A (KU / kg)	2192	23.895	4.2900	24.468	7.2701	3.4444	22	-0.08	106.02	0
106	Vitamin A (KU / kg)	0027	24.275	5.4300	24.468	7.2701	3.4444	22	-0.03	106.02	0
106	Vitamin A (KU / kg)	0676	24.538	0.34300	24.468	7.2701	3.4444	22	0.01	106.02	0
106	Vitamin A (KU / kg)	0905	24.970	5.9460	24.468	7.2701	3.4444	22	0.07	106.02	0
106	Vitamin A (KU / kg)	0689	25.200	0.80000	24.468	7.2701	3.4444	22	0.10	106.02	0
106	Vitamin A (KU / kg)	0910	25.500	6.0000	24.468	7.2701	3.4444	22	0.14	106.02	0
106	Vitamin A (KU / kg)	0610	26.250	4.1000	24.468	7.2701	3.4444	22	0.25	106.02	0
106	Vitamin A (KU / kg)	0563	26.714	4.3540	24.468	7.2701	3.4444	22	0.31	106.02	0
106	Vitamin A (KU / kg)	0148	27.070	0.74000	24.468	7.2701	3.4444	22	0.36	106.02	0
106	Vitamin A (KU / kg)	0861	28.115	11.010	24.468	7.2701	3.4444	22	0.50	106.02	0
106	Vitamin A (KU / kg)	0017	33.703	3.3644	24.468	7.2701	3.4444	22	1.27	106.02	0
106	Vitamin A (KU / kg)	0019	34.550	1.3500	24.468	7.2701	3.4444	22	1.39	106.00	0
106	Vitamin A (KU / kg)	0171	35.160	1.3400	24.468	7.2701	3.4444	22	1.47	106.00	0
106	Vitamin A (KU / kg)	0675	38.585	7.1500	24.468	7.2701	3.4444	22	1.94	106.02	0
107	Vitamin B12 (µg / kg (ppb))	0227	75.750	6.7000				2		107.00	0
107	Vitamin B12 (µg / kg (ppb))	0160	92.500	17.000				2		107.99	0
108	Vitamin D3 (KU / kg)	0208	2.9100	0.84000	3.6590	0.60643	1.4460	5		108.02	0
108	Vitamin D3 (KU / kg)	0227	3.0900	0.82000	3.6590	0.60643	1.4460	5		108.02	0
108	Vitamin D3 (KU / kg)	0675	4.0400	0.28000	3.6590	0.60643	1.4460	5		108.02	0
108	Vitamin D3 (KU / kg)	2192	4.0950	1.3700	3.6590	0.60643	1.4460	5		108.02	0
108	Vitamin D3 (KU / kg)	0861	4.1600	3.9200	3.6590	0.60643	1.4460	5		108.02	0
109	Vitamin E (IU / kg)	2144	91.295	0.63000	119.78	26.100	4.6337	17	-1.09	109.99	0
109	Vitamin E (IU / kg)	0208	93.567	15.124	119.78	26.100	4.6337	17	-1.00	109.02	0
109	Vitamin E (IU / kg)	0098	97.500	7.0000	119.78	26.100	4.6337	17	-0.85	109.02	0
109	Vitamin E (IU / kg)	0941	104.60	2.6000	119.78	26.100	4.6337	17	-0.58	109.02	0
109	Vitamin E (IU / kg)	0590	105.05	11.900	119.78	26.100	4.6337	17	-0.56	109.02	0
109	Vitamin E (IU / kg)	0675	106.23	0.81000	119.78	26.100	4.6337	17	-0.52	109.02	0
109	Vitamin E (IU / kg)	0861	110.00	0.00000	119.78	26.100	4.6337	17	-0.37	109.02	0
109	Vitamin E (IU / kg)	0910	110.40	3.8000	119.78	26.100	4.6337	17	-0.36	109.02	0
109	Vitamin E (IU / kg)	0676	111.11	0.89000	119.78	26.100	4.6337	17	-0.33	109.02	0
109	Vitamin E (IU / kg)	0027	114.44	2.9000	119.78	26.100	4.6337	17	-0.20	109.02	0
109	Vitamin E (IU / kg)	0096	118.29	5.9641	119.78	26.100	4.6337	17	-0.06	109.02	0
109	Vitamin E (IU / kg)	0610	118.50	1.0000	119.78	26.100	4.6337	17	-0.05	109.02	0
109	Vitamin E (IU / kg)	0148	128.70	7.3500	119.78	26.100	4.6337	17	0.34	109.02	0
109	Vitamin E (IU / kg)	0563	150.66	2.2049	119.78	26.100	4.6337	17	1.18	109.02	0
109	Vitamin E (IU / kg)	0905	158.00	6.0000	119.78	26.100	4.6337	17	1.46	109.02	0
109	Vitamin E (IU / kg)	2192	166.60	3.2000	119.78	26.100	4.6337	17	1.79	109.02	0
109	Vitamin E (IU / kg)	2161	274.00	7.4000	119.78	26.100	4.6337	17	5.91	109.02	0
112	Pyridoxine (µg / g)	0160	0.22500	0.03000				2		112.01	0
112	Pyridoxine (µg / g)	0227	3.3300	0.26000				2		112.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
113	Folic Acid (mg / kg (ppm))	0227	1.6500	0.16000				2		113.01	0
113	Folic Acid (mg / kg (ppm))	0160	1.9250	0.03000				2		113.02	0
114	Biotin (mg / kg (ppm))	0160	0.22000	0.08000				2		114.99	0
114	Biotin (mg / kg (ppm))	0227	0.25250	0.00900				2		114.01	0
115	Non Protein N (NPN) (%)	0098	0.33500	0.07000				1		115.00	0
120	Alanine (%)	2161	0.69300	0.01400	0.91821	0.03997	0.01286	27	-5.63	120.99	0
120	Alanine (%)	0148	0.85500	0.01000	0.91821	0.03997	0.01286	27	-1.58	120.05	0
120	Alanine (%)	0870	0.86585	0.00150	0.91821	0.03997	0.01286	27	-1.31	120.00	0
120	Alanine (%)	0859	0.87750	0.01100	0.91821	0.03997	0.01286	27	-1.02	120.00	0
120	Alanine (%)	0610	0.89000	0.00000	0.91821	0.03997	0.01286	27	-0.71	120.05	0
120	Alanine (%)	0626	0.89250	0.01900	0.91821	0.03997	0.01286	27	-0.64	120.05	0
120	Alanine (%)	2146	0.89500	0.01000	0.91821	0.03997	0.01286	27	-0.58	120.99	0
120	Alanine (%)	0226	0.89970	0.04400	0.91821	0.03997	0.01286	27	-0.46	120.00	0
120	Alanine (%)	0910	0.90500	0.05000	0.91821	0.03997	0.01286	27	-0.33	120.00	0
120	Alanine (%)	2059	0.90750	0.00100	0.91821	0.03997	0.01286	27	-0.27	120.00	0
120	Alanine (%)	0644	0.90950	0.00300	0.91821	0.03997	0.01286	27	-0.22	120.00	0
120	Alanine (%)	0652	0.91000	0.00000	0.91821	0.03997	0.01286	27	-0.21	120.00	0
120	Alanine (%)	0968	0.91150	0.01300	0.91821	0.03997	0.01286	27	-0.17	120.00	0
120	Alanine (%)	0872	0.91250	0.00700	0.91821	0.03997	0.01286	27	-0.14	120.00	0
120	Alanine (%)	0571	0.91550	0.00500	0.91821	0.03997	0.01286	27	-0.07	120.00	0
120	Alanine (%)	0098	0.91600	0.02600	0.91821	0.03997	0.01286	27	-0.06	120.02	0
120	Alanine (%)	0675	0.92000	0.00000	0.91821	0.03997	0.01286	27	0.04	120.00	0
120	Alanine (%)	0227	0.92500	0.01000	0.91821	0.03997	0.01286	27	0.17	120.00	0
120	Alanine (%)	0504	0.93000	0.00000	0.91821	0.03997	0.01286	27	0.30	120.00	0
120	Alanine (%)	0878	0.93000	0.00800	0.91821	0.03997	0.01286	27	0.30	120.00	0
120	Alanine (%)	0868	0.93250	0.00700	0.91821	0.03997	0.01286	27	0.36	120.00	0
120	Alanine (%)	2196	0.95900	0.00000	0.91821	0.03997	0.01286	27	1.02	120.05	0
120	Alanine (%)	0353	0.97000	0.00000	0.91821	0.03997	0.01286	27	1.30	120.00	0
120	Alanine (%)	0918	0.97125	0.00470	0.91821	0.03997	0.01286	27	1.33	120.00	0
120	Alanine (%)	0354	0.97300	0.06800	0.91821	0.03997	0.01286	27	1.37	120.00	0
120	Alanine (%)	0676	0.98000	0.02000	0.91821	0.03997	0.01286	27	1.55	120.05	0
120	Alanine (%)	0684	1.1145	0.01500	0.91821	0.03997	0.01286	27	4.91	120.00	0
120	Alanine (%)	2188	1.0500	0.12000	0.91821	0.03997	0.01286	27	3.30	120.05	1
121	Arginine (%)	0148	1.0850	0.09000	1.1823	0.04325	0.03052	27	-2.25	121.05	0
121	Arginine (%)	0870	1.0969	0.01180	1.1823	0.04325	0.03052	27	-1.97	121.00	0
121	Arginine (%)	0859	1.1020	0.00400	1.1823	0.04325	0.03052	27	-1.86	121.00	0
121	Arginine (%)	2196	1.1450	0.00000	1.1823	0.04325	0.03052	27	-0.86	121.05	0
121	Arginine (%)	0626	1.1485	0.05300	1.1823	0.04325	0.03052	27	-0.78	121.05	0
121	Arginine (%)	0353	1.1500	0.00000	1.1823	0.04325	0.03052	27	-0.75	121.00	0
121	Arginine (%)	0878	1.1595	0.00700	1.1823	0.04325	0.03052	27	-0.53	121.00	0
121	Arginine (%)	2188	1.1600	0.04000	1.1823	0.04325	0.03052	27	-0.52	121.05	0
121	Arginine (%)	0227	1.1650	0.11000	1.1823	0.04325	0.03052	27	-0.40	121.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
121	Arginine (%)	0098	1.1660	0.05400	1.1823	0.04325	0.03052	27	-0.38	121.02	0
121	Arginine (%)	0644	1.1745	0.01300	1.1823	0.04325	0.03052	27	-0.18	121.00	0
121	Arginine (%)	0910	1.1750	0.03000	1.1823	0.04325	0.03052	27	-0.17	121.00	0
121	Arginine (%)	0868	1.1765	0.02700	1.1823	0.04325	0.03052	27	-0.13	121.00	0
121	Arginine (%)	0226	1.1792	0.04390	1.1823	0.04325	0.03052	27	-0.07	121.00	0
121	Arginine (%)	0504	1.1850	0.05000	1.1823	0.04325	0.03052	27	0.06	121.00	0
121	Arginine (%)	0652	1.1850	0.01000	1.1823	0.04325	0.03052	27	0.06	121.00	0
121	Arginine (%)	0968	1.1865	0.00100	1.1823	0.04325	0.03052	27	0.10	121.00	0
121	Arginine (%)	2059	1.1900	0.00600	1.1823	0.04325	0.03052	27	0.18	121.00	0
121	Arginine (%)	0675	1.1950	0.01000	1.1823	0.04325	0.03052	27	0.29	121.00	0
121	Arginine (%)	0684	1.2020	0.02600	1.1823	0.04325	0.03052	27	0.46	121.00	0
121	Arginine (%)	0872	1.2095	0.00900	1.1823	0.04325	0.03052	27	0.63	121.00	0
121	Arginine (%)	2161	1.2100	0.01200	1.1823	0.04325	0.03052	27	0.64	121.99	0
121	Arginine (%)	0571	1.2325	0.00900	1.1823	0.04325	0.03052	27	1.16	121.00	0
121	Arginine (%)	0610	1.2350	0.01000	1.1823	0.04325	0.03052	27	1.22	121.05	0
121	Arginine (%)	0918	1.2457	0.00230	1.1823	0.04325	0.03052	27	1.46	121.00	0
121	Arginine (%)	0354	1.2635	0.10500	1.1823	0.04325	0.03052	27	1.88	121.00	0
121	Arginine (%)	0676	1.2650	0.09000	1.1823	0.04325	0.03052	27	1.91	121.05	0
122	Aspartic (%)	0226	1.7104	0.06490	1.8256	0.04250	0.04174	27	-2.71	122.00	0
122	Aspartic (%)	0859	1.7350	0.02000	1.8256	0.04250	0.04174	27	-2.13	122.00	0
122	Aspartic (%)	2146	1.7450	0.07000	1.8256	0.04250	0.04174	27	-1.90	122.99	0
122	Aspartic (%)	0870	1.7695	0.01200	1.8256	0.04250	0.04174	27	-1.32	122.00	0
122	Aspartic (%)	0353	1.7800	0.02000	1.8256	0.04250	0.04174	27	-1.07	122.00	0
122	Aspartic (%)	0868	1.7930	0.00400	1.8256	0.04250	0.04174	27	-0.77	122.00	0
122	Aspartic (%)	0872	1.7970	0.00600	1.8256	0.04250	0.04174	27	-0.67	122.00	0
122	Aspartic (%)	0148	1.8050	0.05000	1.8256	0.04250	0.04174	27	-0.49	122.05	0
122	Aspartic (%)	2059	1.8125	0.02100	1.8256	0.04250	0.04174	27	-0.31	122.00	0
122	Aspartic (%)	0910	1.8150	0.01000	1.8256	0.04250	0.04174	27	-0.25	122.00	0
122	Aspartic (%)	0644	1.8200	0.01000	1.8256	0.04250	0.04174	27	-0.13	122.00	0
122	Aspartic (%)	2188	1.8250	0.17000	1.8256	0.04250	0.04174	27	-0.02	122.05	0
122	Aspartic (%)	0504	1.8250	0.01000	1.8256	0.04250	0.04174	27	-0.02	122.00	0
122	Aspartic (%)	0626	1.8275	0.01500	1.8256	0.04250	0.04174	27	0.04	122.05	0
122	Aspartic (%)	0610	1.8350	0.01000	1.8256	0.04250	0.04174	27	0.22	122.05	0
122	Aspartic (%)	0684	1.8395	0.06900	1.8256	0.04250	0.04174	27	0.33	122.00	0
122	Aspartic (%)	0676	1.8400	0.18000	1.8256	0.04250	0.04174	27	0.34	122.05	0
122	Aspartic (%)	0675	1.8400	0.02000	1.8256	0.04250	0.04174	27	0.34	122.00	0
122	Aspartic (%)	0652	1.8500	0.04000	1.8256	0.04250	0.04174	27	0.57	122.00	0
122	Aspartic (%)	0571	1.8515	0.02300	1.8256	0.04250	0.04174	27	0.61	122.00	0
122	Aspartic (%)	0968	1.8525	0.00500	1.8256	0.04250	0.04174	27	0.63	122.00	0
122	Aspartic (%)	0098	1.8555	0.09300	1.8256	0.04250	0.04174	27	0.70	122.02	0
122	Aspartic (%)	2196	1.8560	0.00000	1.8256	0.04250	0.04174	27	0.71	122.05	0
122	Aspartic (%)	0227	1.8650	0.07000	1.8256	0.04250	0.04174	27	0.93	122.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
122	Aspartic (%)	0878	1.8735	0.01500	1.8256	0.04250	0.04174	27	1.13	122.00	0
122	Aspartic (%)	0354	1.9415	0.07500	1.8256	0.04250	0.04174	27	2.73	122.00	0
122	Aspartic (%)	0918	1.9825	0.04410	1.8256	0.04250	0.04174	27	3.69	122.00	0
122	Aspartic (%)	2161	0.83500	0.01800	1.8256	0.04250	0.04174	27	-23.31	122.99	2
124	Cysteine/Cystine (%)	0353	0.23500	0.03000	0.30091	0.02257	0.01194	28	-2.92	124.00	0
124	Cysteine/Cystine (%)	0870	0.25945	0.00050	0.30091	0.02257	0.01194	28	-1.84	124.00	0
124	Cysteine/Cystine (%)	0859	0.26250	0.00500	0.30091	0.02257	0.01194	28	-1.70	124.00	0
124	Cysteine/Cystine (%)	0148	0.27500	0.03000	0.30091	0.02257	0.01194	28	-1.15	124.05	0
124	Cysteine/Cystine (%)	0227	0.28000	0.00000	0.30091	0.02257	0.01194	28	-0.93	124.00	0
124	Cysteine/Cystine (%)	2196	0.28400	0.00000	0.30091	0.02257	0.01194	28	-0.75	124.05	0
124	Cysteine/Cystine (%)	0652	0.28500	0.01000	0.30091	0.02257	0.01194	28	-0.71	124.00	0
124	Cysteine/Cystine (%)	0918	0.28500	0.01000	0.30091	0.02257	0.01194	28	-0.71	124.00	0
124	Cysteine/Cystine (%)	0644	0.29300	0.01600	0.30091	0.02257	0.01194	28	-0.35	124.00	0
124	Cysteine/Cystine (%)	0868	0.29300	0.01200	0.30091	0.02257	0.01194	28	-0.35	124.00	0
124	Cysteine/Cystine (%)	0872	0.29700	0.02400	0.30091	0.02257	0.01194	28	-0.17	124.00	0
124	Cysteine/Cystine (%)	0968	0.29800	0.00200	0.30091	0.02257	0.01194	28	-0.13	124.00	0
124	Cysteine/Cystine (%)	2059	0.29850	0.00300	0.30091	0.02257	0.01194	28	-0.11	124.00	0
124	Cysteine/Cystine (%)	0675	0.30000	0.00000	0.30091	0.02257	0.01194	28	-0.04	124.00	0
124	Cysteine/Cystine (%)	0684	0.30500	0.00600	0.30091	0.02257	0.01194	28	0.18	124.00	0
124	Cysteine/Cystine (%)	0910	0.30500	0.01000	0.30091	0.02257	0.01194	28	0.18	124.00	0
124	Cysteine/Cystine (%)	2145	0.30500	0.03000	0.30091	0.02257	0.01194	28	0.18	124.99	0
124	Cysteine/Cystine (%)	0098	0.30650	0.00500	0.30091	0.02257	0.01194	28	0.25	124.02	0
124	Cysteine/Cystine (%)	0571	0.30850	0.00500	0.30091	0.02257	0.01194	28	0.34	124.00	0
124	Cysteine/Cystine (%)	0878	0.31100	0.01200	0.30091	0.02257	0.01194	28	0.45	124.00	0
124	Cysteine/Cystine (%)	0504	0.31500	0.01000	0.30091	0.02257	0.01194	28	0.62	124.00	0
124	Cysteine/Cystine (%)	2146	0.31500	0.01000	0.30091	0.02257	0.01194	28	0.62	124.99	0
124	Cysteine/Cystine (%)	0354	0.31600	0.00400	0.30091	0.02257	0.01194	28	0.67	124.00	0
124	Cysteine/Cystine (%)	0226	0.31950	0.00980	0.30091	0.02257	0.01194	28	0.82	124.00	0
124	Cysteine/Cystine (%)	2195	0.32500	0.03000	0.30091	0.02257	0.01194	28	1.07	124.05	0
124	Cysteine/Cystine (%)	0610	0.33500	0.01000	0.30091	0.02257	0.01194	28	1.51	124.05	0
124	Cysteine/Cystine (%)	2188	0.35000	0.02000	0.30091	0.02257	0.01194	28	2.18	124.05	0
124	Cysteine/Cystine (%)	0676	0.37500	0.03000	0.30091	0.02257	0.01194	28	3.28	124.05	0
124	Cysteine/Cystine (%)	2161	1.7450	0.05000	0.30091	0.02257	0.01194	28	64.00	124.99	2
125	Glutamic (%)	2161	2.2500	0.02000	3.3283	0.13552	0.04883	27	-7.96	125.99	0
125	Glutamic (%)	2146	2.8700	0.00000	3.3283	0.13552	0.04883	27	-3.38	125.99	0
125	Glutamic (%)	0859	3.0860	0.03200	3.3283	0.13552	0.04883	27	-1.79	125.00	0
125	Glutamic (%)	2059	3.2110	0.00000	3.3283	0.13552	0.04883	27	-0.87	125.00	0
125	Glutamic (%)	0626	3.2180	0.06400	3.3283	0.13552	0.04883	27	-0.81	125.05	0
125	Glutamic (%)	0910	3.2400	0.04000	3.3283	0.13552	0.04883	27	-0.65	125.00	0
125	Glutamic (%)	0872	3.2820	0.02800	3.3283	0.13552	0.04883	27	-0.34	125.00	0
125	Glutamic (%)	0870	3.2867	0.00220	3.3283	0.13552	0.04883	27	-0.31	125.00	0
125	Glutamic (%)	0644	3.2885	0.00900	3.3283	0.13552	0.04883	27	-0.29	125.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
125	Glutamic (%)	0610	3.2900	0.00000	3.3283	0.13552	0.04883	27	-0.28	125.05	0
125	Glutamic (%)	0504	3.3050	0.05000	3.3283	0.13552	0.04883	27	-0.17	125.00	0
125	Glutamic (%)	0652	3.3100	0.02000	3.3283	0.13552	0.04883	27	-0.13	125.00	0
125	Glutamic (%)	2196	3.3100	0.00000	3.3283	0.13552	0.04883	27	-0.13	125.05	0
125	Glutamic (%)	0571	3.3135	0.04700	3.3283	0.13552	0.04883	27	-0.11	125.00	0
125	Glutamic (%)	0227	3.3150	0.07000	3.3283	0.13552	0.04883	27	-0.10	125.00	0
125	Glutamic (%)	0675	3.3150	0.07000	3.3283	0.13552	0.04883	27	-0.10	125.00	0
125	Glutamic (%)	0868	3.3300	0.03400	3.3283	0.13552	0.04883	27	0.01	125.00	0
125	Glutamic (%)	0878	3.3390	0.03400	3.3283	0.13552	0.04883	27	0.08	125.00	0
125	Glutamic (%)	0098	3.3410	0.14200	3.3283	0.13552	0.04883	27	0.09	125.02	0
125	Glutamic (%)	0968	3.3445	0.01100	3.3283	0.13552	0.04883	27	0.12	125.00	0
125	Glutamic (%)	0148	3.3850	0.09000	3.3283	0.13552	0.04883	27	0.42	125.05	0
125	Glutamic (%)	0226	3.4693	0.10760	3.3283	0.13552	0.04883	27	1.04	125.00	0
125	Glutamic (%)	0684	3.5000	0.13400	3.3283	0.13552	0.04883	27	1.27	125.00	0
125	Glutamic (%)	2188	3.5000	0.04000	3.3283	0.13552	0.04883	27	1.27	125.05	0
125	Glutamic (%)	0354	3.5395	0.24100	3.3283	0.13552	0.04883	27	1.56	125.00	0
125	Glutamic (%)	0918	3.6349	0.00270	3.3283	0.13552	0.04883	27	2.26	125.00	0
125	Glutamic (%)	0353	3.7950	0.03000	3.3283	0.13552	0.04883	27	3.44	125.00	0
125	Glutamic (%)	0676	3.5400	0.28000	3.3283	0.13552	0.04883	27	1.56	125.05	1
126	Glycine (%)	2161	0.55300	0.01400	0.79152	0.03966	0.01877	28	-6.01	126.99	0
126	Glycine (%)	0148	0.69000	0.08000	0.79152	0.03966	0.01877	28	-2.56	126.05	0
126	Glycine (%)	0870	0.72435	0.00090	0.79152	0.03966	0.01877	28	-1.69	126.00	0
126	Glycine (%)	0859	0.74650	0.00700	0.79152	0.03966	0.01877	28	-1.14	126.00	0
126	Glycine (%)	2196	0.75700	0.00000	0.79152	0.03966	0.01877	28	-0.87	126.05	0
126	Glycine (%)	0910	0.76000	0.02000	0.79152	0.03966	0.01877	28	-0.79	126.00	0
126	Glycine (%)	2146	0.76000	0.04000	0.79152	0.03966	0.01877	28	-0.79	126.99	0
126	Glycine (%)	0652	0.77000	0.00000	0.79152	0.03966	0.01877	28	-0.54	126.00	0
126	Glycine (%)	0626	0.78000	0.01600	0.79152	0.03966	0.01877	28	-0.29	126.05	0
126	Glycine (%)	0872	0.78050	0.00700	0.79152	0.03966	0.01877	28	-0.28	126.00	0
126	Glycine (%)	0968	0.78200	0.00400	0.79152	0.03966	0.01877	28	-0.24	126.00	0
126	Glycine (%)	2059	0.78400	0.00400	0.79152	0.03966	0.01877	28	-0.19	126.00	0
126	Glycine (%)	0675	0.78500	0.01000	0.79152	0.03966	0.01877	28	-0.16	126.00	0
126	Glycine (%)	0227	0.79500	0.03000	0.79152	0.03966	0.01877	28	0.09	126.00	0
126	Glycine (%)	0571	0.79550	0.00100	0.79152	0.03966	0.01877	28	0.10	126.00	0
126	Glycine (%)	0644	0.79600	0.00400	0.79152	0.03966	0.01877	28	0.11	126.00	0
126	Glycine (%)	0098	0.79600	0.03800	0.79152	0.03966	0.01877	28	0.11	126.02	0
126	Glycine (%)	0878	0.79800	0.00600	0.79152	0.03966	0.01877	28	0.16	126.00	0
126	Glycine (%)	0504	0.80500	0.01000	0.79152	0.03966	0.01877	28	0.34	126.00	0
126	Glycine (%)	0610	0.80500	0.01000	0.79152	0.03966	0.01877	28	0.34	126.05	0
126	Glycine (%)	0868	0.80700	0.00600	0.79152	0.03966	0.01877	28	0.39	126.00	0
126	Glycine (%)	0684	0.82500	0.02800	0.79152	0.03966	0.01877	28	0.84	126.00	0
126	Glycine (%)	0354	0.82750	0.05100	0.79152	0.03966	0.01877	28	0.91	126.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
126	Glycine (%)	0226	0.83310	0.01440	0.79152	0.03966	0.01877	28	1.05	126.00	0
126	Glycine (%)	0918	0.83635	0.00430	0.79152	0.03966	0.01877	28	1.13	126.00	0
126	Glycine (%)	0353	0.84000	0.02000	0.79152	0.03966	0.01877	28	1.22	126.00	0
126	Glycine (%)	2188	0.91500	0.03000	0.79152	0.03966	0.01877	28	3.11	126.05	0
126	Glycine (%)	0676	1.1650	0.07000	0.79152	0.03966	0.01877	28	9.42	126.05	0
127	Histidine (%)	2188	0.41000	0.00000	0.47963	0.02374	0.01252	28	-2.93	127.05	0
127	Histidine (%)	0859	0.43050	0.00500	0.47963	0.02374	0.01252	28	-2.07	127.00	0
127	Histidine (%)	0870	0.43345	0.00410	0.47963	0.02374	0.01252	28	-1.94	127.00	0
127	Histidine (%)	0148	0.43500	0.05000	0.47963	0.02374	0.01252	28	-1.88	127.05	0
127	Histidine (%)	2161	0.45300	0.01000	0.47963	0.02374	0.01252	28	-1.12	127.99	0
127	Histidine (%)	0626	0.46100	0.00800	0.47963	0.02374	0.01252	28	-0.78	127.05	0
127	Histidine (%)	2059	0.46400	0.00800	0.47963	0.02374	0.01252	28	-0.66	127.00	0
127	Histidine (%)	0910	0.46500	0.01000	0.47963	0.02374	0.01252	28	-0.62	127.00	0
127	Histidine (%)	0878	0.46750	0.01300	0.47963	0.02374	0.01252	28	-0.51	127.00	0
127	Histidine (%)	0098	0.47050	0.01500	0.47963	0.02374	0.01252	28	-0.38	127.02	0
127	Histidine (%)	2196	0.48000	0.00000	0.47963	0.02374	0.01252	28	0.02	127.05	0
127	Histidine (%)	0918	0.48250	0.00960	0.47963	0.02374	0.01252	28	0.12	127.00	0
127	Histidine (%)	0644	0.48300	0.00000	0.47963	0.02374	0.01252	28	0.14	127.00	0
127	Histidine (%)	0227	0.48500	0.01000	0.47963	0.02374	0.01252	28	0.23	127.00	0
127	Histidine (%)	0676	0.48500	0.05000	0.47963	0.02374	0.01252	28	0.23	127.05	0
127	Histidine (%)	2146	0.48500	0.03000	0.47963	0.02374	0.01252	28	0.23	127.99	0
127	Histidine (%)	0872	0.48600	0.01600	0.47963	0.02374	0.01252	28	0.27	127.00	0
127	Histidine (%)	0684	0.48700	0.00400	0.47963	0.02374	0.01252	28	0.31	127.00	0
127	Histidine (%)	0226	0.48925	0.00090	0.47963	0.02374	0.01252	28	0.41	127.00	0
127	Histidine (%)	0675	0.49000	0.02000	0.47963	0.02374	0.01252	28	0.44	127.00	0
127	Histidine (%)	0968	0.49350	0.00100	0.47963	0.02374	0.01252	28	0.58	127.00	0
127	Histidine (%)	0504	0.49500	0.01000	0.47963	0.02374	0.01252	28	0.65	127.00	0
127	Histidine (%)	0868	0.49700	0.00000	0.47963	0.02374	0.01252	28	0.73	127.00	0
127	Histidine (%)	0652	0.50000	0.00000	0.47963	0.02374	0.01252	28	0.86	127.00	0
127	Histidine (%)	0571	0.50150	0.01500	0.47963	0.02374	0.01252	28	0.92	127.00	0
127	Histidine (%)	0610	0.50500	0.01000	0.47963	0.02374	0.01252	28	1.07	127.05	0
127	Histidine (%)	0354	0.51250	0.04100	0.47963	0.02374	0.01252	28	1.38	127.00	0
127	Histidine (%)	0353	0.61500	0.01000	0.47963	0.02374	0.01252	28	5.70	127.00	0
128	Isoleucine (%)	0353	0.58500	0.05000	0.75644	0.04487	0.01976	28	-3.82	128.00	0
128	Isoleucine (%)	0870	0.65765	0.00050	0.75644	0.04487	0.01976	28	-2.20	128.00	0
128	Isoleucine (%)	2161	0.66300	0.01000	0.75644	0.04487	0.01976	28	-2.08	128.99	0
128	Isoleucine (%)	0626	0.69650	0.01300	0.75644	0.04487	0.01976	28	-1.34	128.05	0
128	Isoleucine (%)	0918	0.70725	0.02990	0.75644	0.04487	0.01976	28	-1.10	128.00	0
128	Isoleucine (%)	2146	0.71000	0.04000	0.75644	0.04487	0.01976	28	-1.03	128.99	0
128	Isoleucine (%)	0868	0.73400	0.00800	0.75644	0.04487	0.01976	28	-0.50	128.00	0
128	Isoleucine (%)	0684	0.73850	0.01100	0.75644	0.04487	0.01976	28	-0.40	128.00	0
128	Isoleucine (%)	0098	0.74000	0.03000	0.75644	0.04487	0.01976	28	-0.37	128.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
128	Isoleucine (%)	0859	0.74250	0.00700	0.75644	0.04487	0.01976	28	-0.31	128.00	0
128	Isoleucine (%)	0652	0.74500	0.01000	0.75644	0.04487	0.01976	28	-0.25	128.00	0
128	Isoleucine (%)	0872	0.74750	0.00300	0.75644	0.04487	0.01976	28	-0.20	128.00	0
128	Isoleucine (%)	0675	0.75000	0.02000	0.75644	0.04487	0.01976	28	-0.14	128.00	0
128	Isoleucine (%)	0878	0.75550	0.00900	0.75644	0.04487	0.01976	28	-0.02	128.00	0
128	Isoleucine (%)	2059	0.76300	0.01400	0.75644	0.04487	0.01976	28	0.15	128.00	0
128	Isoleucine (%)	0571	0.76650	0.00100	0.75644	0.04487	0.01976	28	0.22	128.00	0
128	Isoleucine (%)	0910	0.77000	0.06000	0.75644	0.04487	0.01976	28	0.30	128.00	0
128	Isoleucine (%)	0676	0.77000	0.06000	0.75644	0.04487	0.01976	28	0.30	128.05	0
128	Isoleucine (%)	0227	0.78000	0.02000	0.75644	0.04487	0.01976	28	0.53	128.00	0
128	Isoleucine (%)	0644	0.78500	0.01000	0.75644	0.04487	0.01976	28	0.64	128.00	0
128	Isoleucine (%)	0610	0.78500	0.01000	0.75644	0.04487	0.01976	28	0.64	128.05	0
128	Isoleucine (%)	0968	0.78800	0.01600	0.75644	0.04487	0.01976	28	0.70	128.00	0
128	Isoleucine (%)	0354	0.79800	0.08400	0.75644	0.04487	0.01976	28	0.93	128.00	0
128	Isoleucine (%)	2188	0.80000	0.00000	0.75644	0.04487	0.01976	28	0.97	128.05	0
128	Isoleucine (%)	0148	0.80500	0.01000	0.75644	0.04487	0.01976	28	1.08	128.05	0
128	Isoleucine (%)	2196	0.80700	0.00000	0.75644	0.04487	0.01976	28	1.13	128.05	0
128	Isoleucine (%)	0504	0.81000	0.02000	0.75644	0.04487	0.01976	28	1.19	128.00	0
128	Isoleucine (%)	0226	0.81860	0.00680	0.75644	0.04487	0.01976	28	1.39	128.00	0
129	Leucine (%)	2161	1.3810	0.01400	1.5101	0.05002	0.01872	27	-2.58	129.99	0
129	Leucine (%)	0870	1.3943	0.00040	1.5101	0.05002	0.01872	27	-2.31	129.00	0
129	Leucine (%)	2146	1.4300	0.04000	1.5101	0.05002	0.01872	27	-1.60	129.99	0
129	Leucine (%)	0626	1.4455	0.03500	1.5101	0.05002	0.01872	27	-1.29	129.05	0
129	Leucine (%)	0859	1.4735	0.01300	1.5101	0.05002	0.01872	27	-0.73	129.00	0
129	Leucine (%)	0878	1.4780	0.00600	1.5101	0.05002	0.01872	27	-0.64	129.00	0
129	Leucine (%)	0872	1.4790	0.00400	1.5101	0.05002	0.01872	27	-0.62	129.00	0
129	Leucine (%)	2196	1.4920	0.00000	1.5101	0.05002	0.01872	27	-0.36	129.05	0
129	Leucine (%)	0652	1.4950	0.01000	1.5101	0.05002	0.01872	27	-0.30	129.00	0
129	Leucine (%)	0148	1.4950	0.03000	1.5101	0.05002	0.01872	27	-0.30	129.05	0
129	Leucine (%)	2059	1.4970	0.01600	1.5101	0.05002	0.01872	27	-0.26	129.00	0
129	Leucine (%)	0910	1.5000	0.04000	1.5101	0.05002	0.01872	27	-0.20	129.00	0
129	Leucine (%)	0098	1.5040	0.02800	1.5101	0.05002	0.01872	27	-0.12	129.02	0
129	Leucine (%)	0227	1.5050	0.03000	1.5101	0.05002	0.01872	27	-0.10	129.00	0
129	Leucine (%)	0644	1.5100	0.00200	1.5101	0.05002	0.01872	27	0.00	129.00	0
129	Leucine (%)	0675	1.5150	0.01000	1.5101	0.05002	0.01872	27	0.10	129.00	0
129	Leucine (%)	0968	1.5180	0.00400	1.5101	0.05002	0.01872	27	0.16	129.00	0
129	Leucine (%)	0684	1.5265	0.02300	1.5101	0.05002	0.01872	27	0.33	129.00	0
129	Leucine (%)	0571	1.5340	0.00200	1.5101	0.05002	0.01872	27	0.48	129.00	0
129	Leucine (%)	0868	1.5390	0.01000	1.5101	0.05002	0.01872	27	0.58	129.00	0
129	Leucine (%)	0610	1.5400	0.00000	1.5101	0.05002	0.01872	27	0.60	129.05	0
129	Leucine (%)	0353	1.5550	0.01000	1.5101	0.05002	0.01872	27	0.90	129.00	0
129	Leucine (%)	0504	1.5600	0.00000	1.5101	0.05002	0.01872	27	1.00	129.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
129	Leucine (%)	0918	1.5654	0.02270	1.5101	0.05002	0.01872	27	1.10	129.00	0
129	Leucine (%)	0676	1.5700	0.04000	1.5101	0.05002	0.01872	27	1.20	129.05	0
129	Leucine (%)	0226	1.6585	0.06530	1.5101	0.05002	0.01872	27	2.97	129.00	0
129	Leucine (%)	2188	1.6750	0.05000	1.5101	0.05002	0.01872	27	3.30	129.05	0
129	Leucine (%)	0354	1.6110	0.15600	1.5101	0.05002	0.01872	27	2.02	129.00	1
130	L-Lysine (%)	0353	0.95500	0.01000	1.0959	0.05039	0.03248	34	-2.80	130.00	0
130	L-Lysine (%)	0848	0.98500	0.01000	1.0959	0.05039	0.03248	34	-2.20	130.00	0
130	L-Lysine (%)	2053	0.99150	0.01700	1.0959	0.05039	0.03248	34	-2.07	130.99	0
130	L-Lysine (%)	0859	1.0055	0.00700	1.0959	0.05039	0.03248	34	-1.79	130.00	0
130	L-Lysine (%)	0910	1.0450	0.03000	1.0959	0.05039	0.03248	34	-1.01	130.00	0
130	L-Lysine (%)	2195	1.0550	0.05000	1.0959	0.05039	0.03248	34	-0.81	130.05	0
130	L-Lysine (%)	0870	1.0626	0.01400	1.0959	0.05039	0.03248	34	-0.66	130.00	0
130	L-Lysine (%)	0148	1.0700	0.06000	1.0959	0.05039	0.03248	34	-0.51	130.05	0
130	L-Lysine (%)	0652	1.0750	0.01000	1.0959	0.05039	0.03248	34	-0.42	130.00	0
130	L-Lysine (%)	0684	1.0750	0.04200	1.0959	0.05039	0.03248	34	-0.42	130.00	0
130	L-Lysine (%)	0868	1.0770	0.00200	1.0959	0.05039	0.03248	34	-0.38	130.00	0
130	L-Lysine (%)	0968	1.0795	0.00300	1.0959	0.05039	0.03248	34	-0.33	130.00	0
130	L-Lysine (%)	0610	1.0800	0.00000	1.0959	0.05039	0.03248	34	-0.32	130.05	0
130	L-Lysine (%)	0878	1.0815	0.02900	1.0959	0.05039	0.03248	34	-0.29	130.00	0
130	L-Lysine (%)	0626	1.0815	0.03900	1.0959	0.05039	0.03248	34	-0.29	130.05	0
130	L-Lysine (%)	2196	1.0830	0.00000	1.0959	0.05039	0.03248	34	-0.26	130.05	0
130	L-Lysine (%)	0675	1.0850	0.01000	1.0959	0.05039	0.03248	34	-0.22	130.00	0
130	L-Lysine (%)	0571	1.0945	0.02100	1.0959	0.05039	0.03248	34	-0.03	130.00	0
130	L-Lysine (%)	0644	1.0955	0.00300	1.0959	0.05039	0.03248	34	-0.01	130.00	0
130	L-Lysine (%)	2059	1.0955	0.00300	1.0959	0.05039	0.03248	34	-0.01	130.00	0
130	L-Lysine (%)	0918	1.1067	0.01410	1.0959	0.05039	0.03248	34	0.21	130.00	0
130	L-Lysine (%)	0676	1.1100	0.10000	1.0959	0.05039	0.03248	34	0.28	130.05	0
130	L-Lysine (%)	0226	1.1114	0.00220	1.0959	0.05039	0.03248	34	0.31	130.00	0
130	L-Lysine (%)	0227	1.1150	0.05000	1.0959	0.05039	0.03248	34	0.38	130.00	0
130	L-Lysine (%)	2188	1.1250	0.05000	1.0959	0.05039	0.03248	34	0.58	130.05	0
130	L-Lysine (%)	0098	1.1280	0.12200	1.0959	0.05039	0.03248	34	0.64	130.02	0
130	L-Lysine (%)	0504	1.1300	0.04000	1.0959	0.05039	0.03248	34	0.68	130.00	0
130	L-Lysine (%)	0014	1.1300	0.14000	1.0959	0.05039	0.03248	34	0.68	130.05	0
130	L-Lysine (%)	2161	1.1460	0.00800	1.0959	0.05039	0.03248	34	0.99	130.99	0
130	L-Lysine (%)	0354	1.1565	0.07700	1.0959	0.05039	0.03248	34	1.20	130.00	0
130	L-Lysine (%)	0872	1.1750	0.01400	1.0959	0.05039	0.03248	34	1.57	130.00	0
130	L-Lysine (%)	2146	1.2050	0.03000	1.0959	0.05039	0.03248	34	2.16	130.99	0
130	L-Lysine (%)	2145	1.2250	0.01000	1.0959	0.05039	0.03248	34	2.56	130.99	0
130	L-Lysine (%)	0508	1.4219	0.08710	1.0959	0.05039	0.03248	34	6.47	130.99	0
131	Methionine (%)	0148	0.12000	0.02000	0.27226	0.01934	0.01008	30	-7.87	131.05	0
131	Methionine (%)	2161	0.22600	0.00400	0.27226	0.01934	0.01008	30	-2.39	131.99	0
131	Methionine (%)	0626	0.23050	0.01300	0.27226	0.01934	0.01008	30	-2.16	131.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
131	Methionine (%)	0684	0.25150	0.00100	0.27226	0.01934	0.01008	30	-1.07	131.00	0
131	Methionine (%)	0859	0.25200	0.00200	0.27226	0.01934	0.01008	30	-1.05	131.00	0
131	Methionine (%)	0226	0.25355	0.01590	0.27226	0.01934	0.01008	30	-0.97	131.00	0
131	Methionine (%)	0868	0.25900	0.00400	0.27226	0.01934	0.01008	30	-0.69	131.00	0
131	Methionine (%)	0652	0.26000	0.00000	0.27226	0.01934	0.01008	30	-0.63	131.00	0
131	Methionine (%)	0098	0.26100	0.00800	0.27226	0.01934	0.01008	30	-0.58	131.02	0
131	Methionine (%)	0910	0.26500	0.01000	0.27226	0.01934	0.01008	30	-0.38	131.00	0
131	Methionine (%)	2059	0.26900	0.00200	0.27226	0.01934	0.01008	30	-0.17	131.00	0
131	Methionine (%)	0227	0.27000	0.02000	0.27226	0.01934	0.01008	30	-0.12	131.00	0
131	Methionine (%)	0610	0.27000	0.00000	0.27226	0.01934	0.01008	30	-0.12	131.05	0
131	Methionine (%)	0571	0.27200	0.00200	0.27226	0.01934	0.01008	30	-0.01	131.00	0
131	Methionine (%)	0878	0.27200	0.01400	0.27226	0.01934	0.01008	30	-0.01	131.00	0
131	Methionine (%)	0644	0.27400	0.00600	0.27226	0.01934	0.01008	30	0.09	131.00	0
131	Methionine (%)	0968	0.27400	0.00200	0.27226	0.01934	0.01008	30	0.09	131.00	0
131	Methionine (%)	0504	0.27500	0.01000	0.27226	0.01934	0.01008	30	0.14	131.00	0
131	Methionine (%)	2196	0.27600	0.00000	0.27226	0.01934	0.01008	30	0.19	131.05	0
131	Methionine (%)	0872	0.27850	0.00300	0.27226	0.01934	0.01008	30	0.32	131.00	0
131	Methionine (%)	0870	0.27870	0.01160	0.27226	0.01934	0.01008	30	0.33	131.00	0
131	Methionine (%)	0675	0.28000	0.02000	0.27226	0.01934	0.01008	30	0.40	131.00	0
131	Methionine (%)	0918	0.28000	0.02000	0.27226	0.01934	0.01008	30	0.40	131.00	0
131	Methionine (%)	2146	0.28500	0.01000	0.27226	0.01934	0.01008	30	0.66	131.99	0
131	Methionine (%)	2145	0.29000	0.00000	0.27226	0.01934	0.01008	30	0.92	131.99	0
131	Methionine (%)	2188	0.29000	0.02000	0.27226	0.01934	0.01008	30	0.92	131.05	0
131	Methionine (%)	0354	0.29800	0.01400	0.27226	0.01934	0.01008	30	1.33	131.00	0
131	Methionine (%)	2053	0.31500	0.05000	0.27226	0.01934	0.01008	30	2.21	131.99	0
131	Methionine (%)	2195	0.36500	0.01000	0.27226	0.01934	0.01008	30	4.80	131.05	0
131	Methionine (%)	0353	0.37500	0.01000	0.27226	0.01934	0.01008	30	5.31	131.00	0
131	Methionine (%)	0508	0.37035	0.12210	0.27226	0.01934	0.01008	30	5.07	131.99	1
132	Phenylalanine (%)	0870	0.78275	0.01110	0.87899	0.04335	0.02189	28	-2.22	132.00	0
132	Phenylalanine (%)	2161	0.80500	0.00200	0.87899	0.04335	0.02189	28	-1.71	132.99	0
132	Phenylalanine (%)	2188	0.82500	0.05000	0.87899	0.04335	0.02189	28	-1.25	132.05	0
132	Phenylalanine (%)	0626	0.82800	0.02200	0.87899	0.04335	0.02189	28	-1.18	132.05	0
132	Phenylalanine (%)	0148	0.84500	0.09000	0.87899	0.04335	0.02189	28	-0.78	132.05	0
132	Phenylalanine (%)	2146	0.84500	0.05000	0.87899	0.04335	0.02189	28	-0.78	132.99	0
132	Phenylalanine (%)	0859	0.85150	0.00900	0.87899	0.04335	0.02189	28	-0.63	132.00	0
132	Phenylalanine (%)	0878	0.85200	0.00800	0.87899	0.04335	0.02189	28	-0.62	132.00	0
132	Phenylalanine (%)	0872	0.85600	0.01200	0.87899	0.04335	0.02189	28	-0.53	132.00	0
132	Phenylalanine (%)	0684	0.86250	0.03100	0.87899	0.04335	0.02189	28	-0.38	132.00	0
132	Phenylalanine (%)	0098	0.86700	0.02600	0.87899	0.04335	0.02189	28	-0.28	132.02	0
132	Phenylalanine (%)	0675	0.87000	0.04000	0.87899	0.04335	0.02189	28	-0.21	132.00	0
132	Phenylalanine (%)	0227	0.87500	0.03000	0.87899	0.04335	0.02189	28	-0.09	132.00	0
132	Phenylalanine (%)	2196	0.87500	0.00000	0.87899	0.04335	0.02189	28	-0.09	132.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
132	Phenylalanine (%)	0868	0.88300	0.00200	0.87899	0.04335	0.02189	28	0.09	132.00	0
132	Phenylalanine (%)	0652	0.88500	0.01000	0.87899	0.04335	0.02189	28	0.14	132.00	0
132	Phenylalanine (%)	0910	0.88500	0.01000	0.87899	0.04335	0.02189	28	0.14	132.00	0
132	Phenylalanine (%)	0226	0.89035	0.01990	0.87899	0.04335	0.02189	28	0.26	132.00	0
132	Phenylalanine (%)	0644	0.89050	0.00300	0.87899	0.04335	0.02189	28	0.27	132.00	0
132	Phenylalanine (%)	0968	0.89050	0.01100	0.87899	0.04335	0.02189	28	0.27	132.00	0
132	Phenylalanine (%)	0571	0.90400	0.00600	0.87899	0.04335	0.02189	28	0.58	132.00	0
132	Phenylalanine (%)	2059	0.90500	0.00600	0.87899	0.04335	0.02189	28	0.60	132.00	0
132	Phenylalanine (%)	0504	0.91000	0.00000	0.87899	0.04335	0.02189	28	0.72	132.00	0
132	Phenylalanine (%)	0676	0.91500	0.05000	0.87899	0.04335	0.02189	28	0.83	132.05	0
132	Phenylalanine (%)	0354	0.94150	0.04900	0.87899	0.04335	0.02189	28	1.44	132.00	0
132	Phenylalanine (%)	0610	0.95000	0.02000	0.87899	0.04335	0.02189	28	1.64	132.05	0
132	Phenylalanine (%)	0918	0.95065	0.03490	0.87899	0.04335	0.02189	28	1.65	132.00	0
132	Phenylalanine (%)	0353	1.0150	0.01000	0.87899	0.04335	0.02189	28	3.14	132.00	0
133	Proline (%)	0910	0.94000	0.04000	1.0764	0.07246	0.03332	26	-1.88	133.00	0
133	Proline (%)	0870	0.95180	0.01160	1.0764	0.07246	0.03332	26	-1.72	133.00	0
133	Proline (%)	0353	0.97000	0.08000	1.0764	0.07246	0.03332	26	-1.47	133.00	0
133	Proline (%)	0859	0.99850	0.00900	1.0764	0.07246	0.03332	26	-1.08	133.00	0
133	Proline (%)	0868	1.0005	0.02100	1.0764	0.07246	0.03332	26	-1.05	133.00	0
133	Proline (%)	0227	1.0300	0.06000	1.0764	0.07246	0.03332	26	-0.64	133.00	0
133	Proline (%)	2146	1.0400	0.04000	1.0764	0.07246	0.03332	26	-0.50	133.99	0
133	Proline (%)	0626	1.0455	0.04100	1.0764	0.07246	0.03332	26	-0.43	133.05	0
133	Proline (%)	0644	1.0540	0.00800	1.0764	0.07246	0.03332	26	-0.31	133.00	0
133	Proline (%)	0226	1.0631	0.02960	1.0764	0.07246	0.03332	26	-0.18	133.00	0
133	Proline (%)	0652	1.0650	0.07000	1.0764	0.07246	0.03332	26	-0.16	133.00	0
133	Proline (%)	0571	1.0725	0.02100	1.0764	0.07246	0.03332	26	-0.05	133.00	0
133	Proline (%)	0878	1.0745	0.02900	1.0764	0.07246	0.03332	26	-0.03	133.00	0
133	Proline (%)	2059	1.0775	0.00100	1.0764	0.07246	0.03332	26	0.02	133.00	0
133	Proline (%)	0675	1.0850	0.01000	1.0764	0.07246	0.03332	26	0.12	133.00	0
133	Proline (%)	0968	1.0850	0.01000	1.0764	0.07246	0.03332	26	0.12	133.00	0
133	Proline (%)	0148	1.0900	0.06000	1.0764	0.07246	0.03332	26	0.19	133.05	0
133	Proline (%)	0684	1.0985	0.03700	1.0764	0.07246	0.03332	26	0.31	133.00	0
133	Proline (%)	0610	1.1000	0.00000	1.0764	0.07246	0.03332	26	0.33	133.05	0
133	Proline (%)	0872	1.1270	0.01200	1.0764	0.07246	0.03332	26	0.70	133.00	0
133	Proline (%)	0504	1.1400	0.02000	1.0764	0.07246	0.03332	26	0.88	133.00	0
133	Proline (%)	0354	1.1455	0.14900	1.0764	0.07246	0.03332	26	0.95	133.00	0
133	Proline (%)	2196	1.1540	0.00000	1.0764	0.07246	0.03332	26	1.07	133.05	0
133	Proline (%)	2161	1.1710	0.00200	1.0764	0.07246	0.03332	26	1.31	133.99	0
133	Proline (%)	0918	1.1788	0.06500	1.0764	0.07246	0.03332	26	1.41	133.00	0
133	Proline (%)	0676	1.2400	0.04000	1.0764	0.07246	0.03332	26	2.26	133.05	0
133	Proline (%)	2188	1.2800	0.18000	1.0764	0.07246	0.03332	26	2.81	133.05	1
134	Serine (%)	0504	0.75000	0.02000	0.89693	0.04846	0.01553	27	-3.03	134.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
134	Serine (%)	2161	0.75100	0.01000	0.89693	0.04846	0.01553	27	-3.01	134.99	0
134	Serine (%)	0226	0.83040	0.01820	0.89693	0.04846	0.01553	27	-1.37	134.00	0
134	Serine (%)	0859	0.84700	0.00800	0.89693	0.04846	0.01553	27	-1.03	134.00	0
134	Serine (%)	0610	0.85000	0.02000	0.89693	0.04846	0.01553	27	-0.97	134.05	0
134	Serine (%)	0644	0.85850	0.00300	0.89693	0.04846	0.01553	27	-0.79	134.00	0
134	Serine (%)	0910	0.86500	0.03000	0.89693	0.04846	0.01553	27	-0.66	134.00	0
134	Serine (%)	0098	0.87450	0.00100	0.89693	0.04846	0.01553	27	-0.46	134.02	0
134	Serine (%)	0652	0.88000	0.00000	0.89693	0.04846	0.01553	27	-0.35	134.00	0
134	Serine (%)	2059	0.88050	0.00100	0.89693	0.04846	0.01553	27	-0.34	134.00	0
134	Serine (%)	0571	0.89800	0.00200	0.89693	0.04846	0.01553	27	0.02	134.00	0
134	Serine (%)	0878	0.89850	0.00900	0.89693	0.04846	0.01553	27	0.03	134.00	0
134	Serine (%)	0684	0.89900	0.01600	0.89693	0.04846	0.01553	27	0.04	134.00	0
134	Serine (%)	2146	0.90000	0.00000	0.89693	0.04846	0.01553	27	0.06	134.99	0
134	Serine (%)	0626	0.90350	0.03500	0.89693	0.04846	0.01553	27	0.14	134.05	0
134	Serine (%)	0968	0.90400	0.00400	0.89693	0.04846	0.01553	27	0.15	134.00	0
134	Serine (%)	2188	0.90500	0.03000	0.89693	0.04846	0.01553	27	0.17	134.05	0
134	Serine (%)	2196	0.90500	0.00000	0.89693	0.04846	0.01553	27	0.17	134.05	0
134	Serine (%)	0872	0.90650	0.02300	0.89693	0.04846	0.01553	27	0.20	134.00	0
134	Serine (%)	0675	0.91000	0.00000	0.89693	0.04846	0.01553	27	0.27	134.00	0
134	Serine (%)	0227	0.91500	0.03000	0.89693	0.04846	0.01553	27	0.37	134.00	0
134	Serine (%)	0870	0.92275	0.00330	0.89693	0.04846	0.01553	27	0.53	134.00	0
134	Serine (%)	0868	0.93700	0.01600	0.89693	0.04846	0.01553	27	0.83	134.00	0
134	Serine (%)	0918	0.98100	0.04880	0.89693	0.04846	0.01553	27	1.73	134.00	0
134	Serine (%)	0676	0.98500	0.05000	0.89693	0.04846	0.01553	27	1.82	134.05	0
134	Serine (%)	0354	0.98650	0.03100	0.89693	0.04846	0.01553	27	1.85	134.00	0
134	Serine (%)	0353	1.0050	0.01000	0.89693	0.04846	0.01553	27	2.23	134.00	0
134	Serine (%)	0148	0.83000	0.08000	0.89693	0.04846	0.01553	27	-1.38	134.05	1
135	Threonine (%)	0226	0.65315	0.01330	0.70006	0.02793	0.01566	29	-1.68	135.00	0
135	Threonine (%)	0859	0.65400	0.00800	0.70006	0.02793	0.01566	29	-1.65	135.00	0
135	Threonine (%)	0626	0.65700	0.02400	0.70006	0.02793	0.01566	29	-1.54	135.05	0
135	Threonine (%)	0148	0.67000	0.04000	0.70006	0.02793	0.01566	29	-1.08	135.05	0
135	Threonine (%)	2146	0.67500	0.01000	0.70006	0.02793	0.01566	29	-0.90	135.99	0
135	Threonine (%)	2161	0.67900	0.01400	0.70006	0.02793	0.01566	29	-0.75	135.99	0
135	Threonine (%)	0504	0.68500	0.01000	0.70006	0.02793	0.01566	29	-0.54	135.00	0
135	Threonine (%)	0910	0.68500	0.03000	0.70006	0.02793	0.01566	29	-0.54	135.00	0
135	Threonine (%)	2059	0.68800	0.00400	0.70006	0.02793	0.01566	29	-0.43	135.00	0
135	Threonine (%)	0652	0.69000	0.00000	0.70006	0.02793	0.01566	29	-0.36	135.00	0
135	Threonine (%)	0644	0.69450	0.00300	0.70006	0.02793	0.01566	29	-0.20	135.00	0
135	Threonine (%)	0676	0.69500	0.05000	0.70006	0.02793	0.01566	29	-0.18	135.05	0
135	Threonine (%)	2145	0.69500	0.03000	0.70006	0.02793	0.01566	29	-0.18	135.99	0
135	Threonine (%)	0872	0.69650	0.00700	0.70006	0.02793	0.01566	29	-0.13	135.00	0
135	Threonine (%)	0870	0.69835	0.00510	0.70006	0.02793	0.01566	29	-0.06	135.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
135	Threonine (%)	0878	0.70000	0.00600	0.70006	0.02793	0.01566	29	0.00	135.00	0
135	Threonine (%)	0610	0.70000	0.00000	0.70006	0.02793	0.01566	29	0.00	135.05	0
135	Threonine (%)	0571	0.70350	0.00900	0.70006	0.02793	0.01566	29	0.12	135.00	0
135	Threonine (%)	0098	0.70400	0.02400	0.70006	0.02793	0.01566	29	0.14	135.02	0
135	Threonine (%)	0968	0.71050	0.00100	0.70006	0.02793	0.01566	29	0.37	135.00	0
135	Threonine (%)	0868	0.71250	0.00500	0.70006	0.02793	0.01566	29	0.45	135.00	0
135	Threonine (%)	0227	0.71500	0.01000	0.70006	0.02793	0.01566	29	0.54	135.00	0
135	Threonine (%)	0675	0.71500	0.01000	0.70006	0.02793	0.01566	29	0.54	135.00	0
135	Threonine (%)	2196	0.71800	0.00000	0.70006	0.02793	0.01566	29	0.64	135.05	0
135	Threonine (%)	0684	0.72950	0.02700	0.70006	0.02793	0.01566	29	1.05	135.00	0
135	Threonine (%)	0353	0.75500	0.01000	0.70006	0.02793	0.01566	29	1.97	135.00	0
135	Threonine (%)	0354	0.77250	0.01500	0.70006	0.02793	0.01566	29	2.59	135.00	0
135	Threonine (%)	2188	0.78000	0.06000	0.70006	0.02793	0.01566	29	2.86	135.05	0
135	Threonine (%)	0918	0.78110	0.02860	0.70006	0.02793	0.01566	29	2.90	135.00	0
136	Tryptophan (%)	2161	0.15050	0.00100	0.21899	0.02406	0.00497	16	-2.85	136.99	0
136	Tryptophan (%)	0870	0.17705	0.00050	0.21899	0.02406	0.00497	16	-1.74	136.00	0
136	Tryptophan (%)	0910	0.19500	0.01000	0.21899	0.02406	0.00497	16	-1.00	136.00	0
136	Tryptophan (%)	0918	0.20000	0.00000	0.21899	0.02406	0.00497	16	-0.79	136.00	0
136	Tryptophan (%)	0504	0.20250	0.00100	0.21899	0.02406	0.00497	16	-0.69	136.99	0
136	Tryptophan (%)	0868	0.21100	0.00400	0.21899	0.02406	0.00497	16	-0.33	136.01	0
136	Tryptophan (%)	0098	0.21200	0.00000	0.21899	0.02406	0.00497	16	-0.29	136.02	0
136	Tryptophan (%)	2059	0.22400	0.00400	0.21899	0.02406	0.00497	16	0.21	136.03	0
136	Tryptophan (%)	0872	0.22500	0.00400	0.21899	0.02406	0.00497	16	0.25	136.03	0
136	Tryptophan (%)	0684	0.23350	0.02100	0.21899	0.02406	0.00497	16	0.60	136.00	0
136	Tryptophan (%)	0878	0.23450	0.00100	0.21899	0.02406	0.00497	16	0.64	136.01	0
136	Tryptophan (%)	0571	0.23500	0.00200	0.21899	0.02406	0.00497	16	0.67	136.03	0
136	Tryptophan (%)	0859	0.23700	0.01000	0.21899	0.02406	0.00497	16	0.75	136.03	0
136	Tryptophan (%)	0644	0.23850	0.00100	0.21899	0.02406	0.00497	16	0.81	136.01	0
136	Tryptophan (%)	0227	0.24500	0.01000	0.21899	0.02406	0.00497	16	1.08	136.00	0
136	Tryptophan (%)	0610	0.24500	0.01000	0.21899	0.02406	0.00497	16	1.08	136.99	0
137	Tyrosine (%)	0910	0.36500	0.01000	0.59982	0.09479	0.03080	22	-2.48	137.00	0
137	Tyrosine (%)	0870	0.44810	0.00780	0.59982	0.09479	0.03080	22	-1.60	137.02	0
137	Tyrosine (%)	2161	0.47300	0.01800	0.59982	0.09479	0.03080	22	-1.34	137.99	0
137	Tyrosine (%)	2146	0.49500	0.05000	0.59982	0.09479	0.03080	22	-1.11	137.99	0
137	Tyrosine (%)	0626	0.51300	0.01000	0.59982	0.09479	0.03080	22	-0.92	137.05	0
137	Tyrosine (%)	0227	0.56500	0.05000	0.59982	0.09479	0.03080	22	-0.37	137.00	0
137	Tyrosine (%)	0675	0.56500	0.01000	0.59982	0.09479	0.03080	22	-0.37	137.00	0
137	Tyrosine (%)	0968	0.57650	0.01300	0.59982	0.09479	0.03080	22	-0.25	137.00	0
137	Tyrosine (%)	2196	0.59600	0.00000	0.59982	0.09479	0.03080	22	-0.04	137.05	0
137	Tyrosine (%)	0504	0.60000	0.04000	0.59982	0.09479	0.03080	22	0.00	137.00	0
137	Tyrosine (%)	0098	0.60350	0.02900	0.59982	0.09479	0.03080	22	0.04	137.02	0
137	Tyrosine (%)	0918	0.60570	0.00580	0.59982	0.09479	0.03080	22	0.06	137.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
137	Tyrosine (%)	0872	0.60800	0.00200	0.59982	0.09479	0.03080	22	0.09	137.00	0
137	Tyrosine (%)	0226	0.61355	0.04290	0.59982	0.09479	0.03080	22	0.14	137.00	0
137	Tyrosine (%)	2059	0.63250	0.00500	0.59982	0.09479	0.03080	22	0.34	137.00	0
137	Tyrosine (%)	0644	0.63800	0.01000	0.59982	0.09479	0.03080	22	0.40	137.00	0
137	Tyrosine (%)	0354	0.63900	0.10400	0.59982	0.09479	0.03080	22	0.41	137.00	0
137	Tyrosine (%)	0610	0.68500	0.01000	0.59982	0.09479	0.03080	22	0.90	137.05	0
137	Tyrosine (%)	0148	0.69500	0.13000	0.59982	0.09479	0.03080	22	1.00	137.05	0
137	Tyrosine (%)	0676	0.71000	0.04000	0.59982	0.09479	0.03080	22	1.16	137.05	0
137	Tyrosine (%)	2188	0.72500	0.07000	0.59982	0.09479	0.03080	22	1.32	137.05	0
137	Tyrosine (%)	0353	0.82000	0.02000	0.59982	0.09479	0.03080	22	2.32	137.99	0
138	Valine (%)	2161	0.72000	0.00400	0.85883	0.04991	0.01594	27	-2.78	138.99	0
138	Valine (%)	0870	0.74530	0.01160	0.85883	0.04991	0.01594	27	-2.27	138.00	0
138	Valine (%)	0353	0.76000	0.02000	0.85883	0.04991	0.01594	27	-1.98	138.00	0
138	Valine (%)	0626	0.79250	0.01700	0.85883	0.04991	0.01594	27	-1.33	138.05	0
138	Valine (%)	0859	0.81050	0.01500	0.85883	0.04991	0.01594	27	-0.97	138.00	0
138	Valine (%)	2188	0.82000	0.00000	0.85883	0.04991	0.01594	27	-0.78	138.05	0
138	Valine (%)	2146	0.83000	0.02000	0.85883	0.04991	0.01594	27	-0.58	138.99	0
138	Valine (%)	0652	0.83500	0.01000	0.85883	0.04991	0.01594	27	-0.48	138.00	0
138	Valine (%)	0148	0.84500	0.05000	0.85883	0.04991	0.01594	27	-0.28	138.05	0
138	Valine (%)	0684	0.84750	0.00500	0.85883	0.04991	0.01594	27	-0.23	138.00	0
138	Valine (%)	2196	0.84900	0.00000	0.85883	0.04991	0.01594	27	-0.20	138.05	0
138	Valine (%)	2059	0.85400	0.01000	0.85883	0.04991	0.01594	27	-0.10	138.00	0
138	Valine (%)	0675	0.85500	0.01000	0.85883	0.04991	0.01594	27	-0.08	138.00	0
138	Valine (%)	0868	0.85550	0.02900	0.85883	0.04991	0.01594	27	-0.07	138.00	0
138	Valine (%)	0872	0.86100	0.01000	0.85883	0.04991	0.01594	27	0.04	138.00	0
138	Valine (%)	0878	0.86700	0.00400	0.85883	0.04991	0.01594	27	0.16	138.00	0
138	Valine (%)	0098	0.86800	0.02600	0.85883	0.04991	0.01594	27	0.18	138.02	0
138	Valine (%)	0571	0.87750	0.01300	0.85883	0.04991	0.01594	27	0.37	138.00	0
138	Valine (%)	0227	0.89000	0.02000	0.85883	0.04991	0.01594	27	0.62	138.00	0
138	Valine (%)	0644	0.89150	0.00700	0.85883	0.04991	0.01594	27	0.65	138.00	0
138	Valine (%)	0968	0.89500	0.00200	0.85883	0.04991	0.01594	27	0.72	138.00	0
138	Valine (%)	0226	0.90375	0.01890	0.85883	0.04991	0.01594	27	0.90	138.00	0
138	Valine (%)	0676	0.91000	0.06000	0.85883	0.04991	0.01594	27	1.03	138.05	0
138	Valine (%)	0504	0.91000	0.02000	0.85883	0.04991	0.01594	27	1.03	138.00	0
138	Valine (%)	0918	0.91980	0.00780	0.85883	0.04991	0.01594	27	1.22	138.00	0
138	Valine (%)	0354	0.92400	0.03000	0.85883	0.04991	0.01594	27	1.31	138.00	0
138	Valine (%)	0610	0.92500	0.01000	0.85883	0.04991	0.01594	27	1.33	138.05	0
138	Valine (%)	0910	0.88000	0.08000	0.85883	0.04991	0.01594	27	0.42	138.00	1
139	Taurine (%)	0504	0.16000	0.02000				1		139.00	0
139	Taurine (%)	0098	< 0.01					1		139.02	5
139	Taurine (%)	2161	< 0.01					1		139.99	5
150	Phytase (Units / kg)	0047	1,513.5	69.000	2,520.3	1,370.6	87.500	4		150.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
150	Phytase (Units / kg)	0610	1,865.0	30.000	2,520.3	1,370.6	87.500	4	150.00	0	
150	Phytase (Units / kg)	0027	2,165.5	207.00	2,520.3	1,370.6	87.500	4	150.00	0	
150	Phytase (Units / kg)	0939	4,537.0	44.000	2,520.3	1,370.6	87.500	4	150.99	0	
160	Fructose (%)	0096	0.19440	0.01680				3	160.99	0	
160	Fructose (%)	0861	0.20000	0.00000				3	160.99	0	
160	Fructose (%)	0227	0.25500	0.01000				3	160.99	0	
160	Fructose (%)	0148	0.45000	0.08000				3	160.99	1	
162	Glucose (%)	0861	0.15000	0.10000	0.23001	0.07131	0.08378	4	162.99	0	
162	Glucose (%)	0148	0.19000	0.02000	0.23001	0.07131	0.08378	4	162.99	0	
162	Glucose (%)	0227	0.28500	0.11000	0.23001	0.07131	0.08378	4	162.99	0	
162	Glucose (%)	0096	0.29505	0.10510	0.23001	0.07131	0.08378	4	162.99	0	
163	Lactose (%)	0096	1.1086	0.09690	1.1246	0.01954	0.06923	4	163.99	0	
163	Lactose (%)	0227	1.1100	0.08000	1.1246	0.01954	0.06923	4	163.99	0	
163	Lactose (%)	0148	1.1300	0.00000	1.1246	0.01954	0.06923	4	163.99	0	
163	Lactose (%)	0629	1.1500	0.10000	1.1246	0.01954	0.06923	4	163.99	0	
163	Lactose (%)	0861	1.3500	0.10000	1.1246	0.01954	0.06923	4	163.99	2	
164	Maltose (%)	0861	0.10000	0.00000				3	164.99	0	
164	Maltose (%)	0629	0.62500	0.05000				3	164.99	0	
164	Maltose (%)	0096	0.88800	0.39020				3	164.99	0	
164	Maltose (%)	0148	0.00000	0.00000				3	164.99	4	
164	Maltose (%)	0227	< 0.15					3	164.99	5	
165	Sucrose (%)	0227	2.6350	0.09000	2.9163	0.24904	0.08250	4	165.99	0	
165	Sucrose (%)	0629	2.8000	0.00000	2.9163	0.24904	0.08250	4	165.99	0	
165	Sucrose (%)	0148	3.0300	0.04000	2.9163	0.24904	0.08250	4	165.99	0	
165	Sucrose (%)	0861	3.2000	0.20000	2.9163	0.24904	0.08250	4	165.99	0	
165	Sucrose (%)	0096	3.0997	0.70470	2.9163	0.24904	0.08250	4	165.99	1	
166	Raffinose (%)	0227	0.43000	0.04000				1	166.99	0	
167	Stachyose (%)	0227	1.3600	0.10000				1	167.99	0	
348	Bacitracin (mg/kg (ppm))	0964	168.69	30.435				1	348.01	0	
350	Carbadox (mg/kg (ppm))	0047	19.600	0.60000	22.675	2.8221	0.60000	4	350.01	0	
350	Carbadox (mg/kg (ppm))	0001	21.550	0.90000	22.675	2.8221	0.60000	4	350.01	0	
350	Carbadox (mg/kg (ppm))	0036	23.300	0.80000	22.675	2.8221	0.60000	4	350.01	0	
350	Carbadox (mg/kg (ppm))	2053	26.250	0.10000	22.675	2.8221	0.60000	4	350.02	0	
350	Carbadox (mg/kg (ppm))	0038	29.880	11.760	22.675	2.8221	0.60000	4	350.01	1	
351	Chlortetracycline (mg/kg (ppm))	2127	37.250	3.5000	69.975	15.981	2.8929	20	-2.05	351.05	0
351	Chlortetracycline (mg/kg (ppm))	2145	48.075	0.93000	69.975	15.981	2.8929	20	-1.37	351.05	0
351	Chlortetracycline (mg/kg (ppm))	2144	51.250	2.9800	69.975	15.981	2.8929	20	-1.17	351.03	0
351	Chlortetracycline (mg/kg (ppm))	0910	55.500	1.0000	69.975	15.981	2.8929	20	-0.91	351.03	0
351	Chlortetracycline (mg/kg (ppm))	0035	58.465	1.4300	69.975	15.981	2.8929	20	-0.72	351.00	0
351	Chlortetracycline (mg/kg (ppm))	0032	62.000	2.0000	69.975	15.981	2.8929	20	-0.50	351.03	0
351	Chlortetracycline (mg/kg (ppm))	2053	62.500	3.0000	69.975	15.981	2.8929	20	-0.47	351.04	0
351	Chlortetracycline (mg/kg (ppm))	0009	64.175	0.95000	69.975	15.981	2.8929	20	-0.36	351.03	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
351	Chlortetracycline (mg/kg (ppm))	0917	68.000	6.0000	69.975	15.981	2.8929	20	-0.12	351.03	0
351	Chlortetracycline (mg/kg (ppm))	0043	71.150	2.3000	69.975	15.981	2.8929	20	0.07	351.00	0
351	Chlortetracycline (mg/kg (ppm))	0964	71.992	0.02210	69.975	15.981	2.8929	20	0.13	351.00	0
351	Chlortetracycline (mg/kg (ppm))	0218	73.512	2.4067	69.975	15.981	2.8929	20	0.22	351.00	0
351	Chlortetracycline (mg/kg (ppm))	0038	73.775	0.35000	69.975	15.981	2.8929	20	0.24	351.03	0
351	Chlortetracycline (mg/kg (ppm))	0028	76.450	3.9000	69.975	15.981	2.8929	20	0.41	351.05	0
351	Chlortetracycline (mg/kg (ppm))	0036	77.550	1.1000	69.975	15.981	2.8929	20	0.47	351.03	0
351	Chlortetracycline (mg/kg (ppm))	0027	80.605	3.4900	69.975	15.981	2.8929	20	0.67	351.00	0
351	Chlortetracycline (mg/kg (ppm))	0001	82.150	2.7000	69.975	15.981	2.8929	20	0.76	351.03	0
351	Chlortetracycline (mg/kg (ppm))	0010	90.450	5.9000	69.975	15.981	2.8929	20	1.28	351.03	0
351	Chlortetracycline (mg/kg (ppm))	0033	91.950	0.90000	69.975	15.981	2.8929	20	1.38	351.03	0
351	Chlortetracycline (mg/kg (ppm))	0511	111.50	13.000	69.975	15.981	2.8929	20	2.60	351.00	0
351	Chlortetracycline (mg/kg (ppm))	0160	86.550	16.700	69.975	15.981	2.8929	20	1.04	351.05	1
357	Ethoxyquin (mg/kg (ppm))	0939	0.32000	0.08000				1		357.01	0
377	Pyrantel Tartrate (mg/kg (ppm))	0038	37.985	0.23000				3		377.01	0
377	Pyrantel Tartrate (mg/kg (ppm))	2053	56.500	1.0000				3		377.02	0
377	Pyrantel Tartrate (mg/kg (ppm))	0047	94.650	8.5000				3		377.01	0
382	Sulfamethazine (mg/kg (ppm))	2053	57.500	2.4000	78.934	18.905	1.9760	5		382.03	0
382	Sulfamethazine (mg/kg (ppm))	0035	70.070	0.88000	78.934	18.905	1.9760	5		382.00	0
382	Sulfamethazine (mg/kg (ppm))	0218	78.050	1.5000	78.934	18.905	1.9760	5		382.02	0
382	Sulfamethazine (mg/kg (ppm))	0001	80.300	4.8000	78.934	18.905	1.9760	5		382.01	0
382	Sulfamethazine (mg/kg (ppm))	0019	108.75	0.30000	78.934	18.905	1.9760	5		382.01	0
382	Sulfamethazine (mg/kg (ppm))	0038	128.05	70.300	78.934	18.905	1.9760	5		382.03	1
386	Tiamulin (mg/kg (ppm))	2144	101.40	4.6000	144.19	29.860	2.9071	7	-1.43	386.02	0
386	Tiamulin (mg/kg (ppm))	0218	122.63	2.2300	144.19	29.860	2.9071	7	-0.72	386.00	0
386	Tiamulin (mg/kg (ppm))	0668	133.88	1.4300	144.19	29.860	2.9071	7	-0.35	386.00	0
386	Tiamulin (mg/kg (ppm))	2053	150.50	1.0000	144.19	29.860	2.9071	7	0.21	386.01	0
386	Tiamulin (mg/kg (ppm))	0001	158.00	2.0000	144.19	29.860	2.9071	7	0.46	386.00	0
386	Tiamulin (mg/kg (ppm))	0910	167.05	8.9000	144.19	29.860	2.9071	7	0.77	386.00	0
386	Tiamulin (mg/kg (ppm))	0875	175.89	0.19000	144.19	29.860	2.9071	7	1.06	386.02	0
400	Water Activity (Units)	2073	0.44800	0.00400	0.49736	0.02737	0.00430	9	-1.80	400.99	0
400	Water Activity (Units)	0843	0.48550	0.00100	0.49736	0.02737	0.00430	9	-0.43	400.99	0
400	Water Activity (Units)	0843	0.48750	0.00760	0.49736	0.02737	0.00430	9	-0.36	400.01	0
400	Water Activity (Units)	0942	0.48850	0.00700	0.49736	0.02737	0.00430	9	-0.32	400.01	0
400	Water Activity (Units)	0407	0.48905	0.00270	0.49736	0.02737	0.00430	9	-0.30	400.01	0
400	Water Activity (Units)	0589	0.49100	0.01000	0.49736	0.02737	0.00430	9	-0.23	400.01	0
400	Water Activity (Units)	2109	0.51950	0.00020	0.49736	0.02737	0.00430	9	0.81	400.01	0
400	Water Activity (Units)	2181	0.52940	0.00520	0.49736	0.02737	0.00430	9	1.17	400.01	0
400	Water Activity (Units)	2009	0.52950	0.00100	0.49736	0.02737	0.00430	9	1.17	400.01	0
412	Starch, Dietary (%)	0353	30.040	1.0000				1		412.01	0
516	Arsenic, Total (mg / kg (ppm))	0910	0.41500	0.07000	0.46885	0.06132	0.02293	8	-0.88	516.52	0
516	Arsenic, Total (mg / kg (ppm))	0045	0.41900	0.03000	0.46885	0.06132	0.02293	8	-0.81	516.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
516	Arsenic, Total (mg / kg (ppm))	0171	0.42650	0.02300	0.46885	0.06132	0.02293	8	-0.69	516.00	0
516	Arsenic, Total (mg / kg (ppm))	0186	0.44850	0.00300	0.46885	0.06132	0.02293	8	-0.33	516.52	0
516	Arsenic, Total (mg / kg (ppm))	0227	0.46750	0.00900	0.46885	0.06132	0.02293	8	-0.02	516.53	0
516	Arsenic, Total (mg / kg (ppm))	0553	0.47900	0.00600	0.46885	0.06132	0.02293	8	0.17	516.53	0
516	Arsenic, Total (mg / kg (ppm))	0098	0.53450	0.01700	0.46885	0.06132	0.02293	8	1.07	516.53	0
516	Arsenic, Total (mg / kg (ppm))	0918	0.58800	0.02540	0.46885	0.06132	0.02293	8	1.94	516.53	0
518	Cadmium (mg / kg (ppm))	0910	0.23000	0.04000	0.27698	0.03460	0.02219	7	-1.36	518.52	0
518	Cadmium (mg / kg (ppm))	0918	0.26535	0.00230	0.27698	0.03460	0.02219	7	-0.34	518.53	0
518	Cadmium (mg / kg (ppm))	0186	0.26750	0.01700	0.27698	0.03460	0.02219	7	-0.27	518.52	0
518	Cadmium (mg / kg (ppm))	0553	0.26900	0.02200	0.27698	0.03460	0.02219	7	-0.23	518.53	0
518	Cadmium (mg / kg (ppm))	0098	0.27700	0.00800	0.27698	0.03460	0.02219	7	0.00	518.53	0
518	Cadmium (mg / kg (ppm))	0227	0.30500	0.06200	0.27698	0.03460	0.02219	7	0.81	518.53	0
518	Cadmium (mg / kg (ppm))	0171	0.32500	0.00400	0.27698	0.03460	0.02219	7	1.39	518.41	0
520	Chromium (mg / kg (ppm))	0171	4.5100	0.03600	6.8160	1.8165	0.68317	9	-1.27	520.41	0
520	Chromium (mg / kg (ppm))	0186	4.6250	0.09000	6.8160	1.8165	0.68317	9	-1.21	520.52	0
520	Chromium (mg / kg (ppm))	0918	5.7729	0.86350	6.8160	1.8165	0.68317	9	-0.57	520.53	0
520	Chromium (mg / kg (ppm))	0693	6.0500	0.10000	6.8160	1.8165	0.68317	9	-0.42	520.42	0
520	Chromium (mg / kg (ppm))	0553	7.5050	2.4700	6.8160	1.8165	0.68317	9	0.38	520.53	0
520	Chromium (mg / kg (ppm))	0045	7.8000	2.0400	6.8160	1.8165	0.68317	9	0.54	520.42	0
520	Chromium (mg / kg (ppm))	0032	8.0765	0.27900	6.8160	1.8165	0.68317	9	0.69	520.53	0
520	Chromium (mg / kg (ppm))	0510	8.4200	0.02000	6.8160	1.8165	0.68317	9	0.88	520.43	0
520	Chromium (mg / kg (ppm))	0955	8.5850	0.25000	6.8160	1.8165	0.68317	9	0.97	520.43	0
526	Lead (mg / kg (ppm))	0910	0.29500	0.01000	0.33243	0.03074	0.00759	7	-1.22	526.52	0
526	Lead (mg / kg (ppm))	0553	0.30650	0.00100	0.33243	0.03074	0.00759	7	-0.84	526.53	0
526	Lead (mg / kg (ppm))	0918	0.33145	0.01810	0.33243	0.03074	0.00759	7	-0.03	526.53	0
526	Lead (mg / kg (ppm))	0098	0.33150	0.00700	0.33243	0.03074	0.00759	7	-0.03	526.53	0
526	Lead (mg / kg (ppm))	0186	0.33850	0.00300	0.33243	0.03074	0.00759	7	0.20	526.52	0
526	Lead (mg / kg (ppm))	0227	0.34550	0.01100	0.33243	0.03074	0.00759	7	0.43	526.53	0
526	Lead (mg / kg (ppm))	0171	0.71050	0.00300	0.33243	0.03074	0.00759	7	12.30	526.41	0
526	Lead (mg / kg (ppm))	0032	0.42050	0.07500	0.33243	0.03074	0.00759	7	2.87	526.53	1
539	Nickel (mg / kg (ppm))	0186	3.0800	0.02000	3.6196	0.60798	0.15125	4		539.52	0
539	Nickel (mg / kg (ppm))	0171	3.1430	0.00800	3.6196	0.60798	0.15125	4		539.41	0
539	Nickel (mg / kg (ppm))	0553	3.9350	0.31000	3.6196	0.60798	0.15125	4		539.53	0
539	Nickel (mg / kg (ppm))	0918	4.3204	0.26700	3.6196	0.60798	0.15125	4		539.53	0
702	Butyric Acid (4:0) (%)	2161	< 0.01					0		702.00	5
703	Valeric Acid (5:0) (%)	2161	0.17000	0.00000				1		703.00	0
704	Caproic Acid (6:0) (%)	2161	0.25000	0.00000				1		704.00	0
706	Caprylic acid (8:0) (%) (w/w)	2161	0.00210	0.00040				1		706.99	0
708	Capric acid (10:0) (%) (w/w)	2161	0.00145	0.00010				1		708.01	0
710	Lauric Acid (12:0) (%) (w/w)	2161	0.00235	0.00010				3		710.99	0
710	Lauric Acid (12:0) (%) (w/w)	0226	0.00285	0.00030				3		710.99	0
710	Lauric Acid (12:0) (%) (w/w)	0164	0.01000	0.00000				3		710.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
714	Myristic Acid (14:0) (% (w/w))	0226	0.01780	0.00120				3	714.99	0	
714	Myristic Acid (14:0) (% (w/w))	0164	0.02500	0.01000				3	714.99	0	
714	Myristic Acid (14:0) (% (w/w))	2161	0.02500	0.00200				3	714.99	0	
716	Palmitic Acid (16:0) (% (w/w))	0226	0.64425	0.05210				3	716.99	0	
716	Palmitic Acid (16:0) (% (w/w))	0164	0.71500	0.05000				3	716.99	0	
716	Palmitic Acid (16:0) (% (w/w))	2161	0.81150	0.06520				3	716.99	0	
718	Palmitoleic Acid (9c-16:1) (% (w/w))	0226	0.02440	0.00120	0.03366	0.00935	0.00778	4	718.99	0	
718	Palmitoleic Acid (9c-16:1) (% (w/w))	0676	0.02700	0.00600	0.03366	0.00935	0.00778	4	718.99	0	
718	Palmitoleic Acid (9c-16:1) (% (w/w))	0164	0.04000	0.02000	0.03366	0.00935	0.00778	4	718.99	0	
718	Palmitoleic Acid (9c-16:1) (% (w/w))	2161	0.04325	0.00390	0.03366	0.00935	0.00778	4	718.99	0	
720	Margaric acid (17:0) (% (w/w))	2161	0.00745	0.00090				1	720.99	0	
722	Stearic Acid (18:0) (% (w/w))	0226	0.15560	0.01140				3	722.99	0	
722	Stearic Acid (18:0) (% (w/w))	0164	0.18000	0.02000				3	722.99	0	
722	Stearic Acid (18:0) (% (w/w))	2161	0.20945	0.01690				3	722.99	0	
724	Oleic Acid (9c-18:1) (% (w/w))	0226	0.94420	0.06440				3	724.99	0	
724	Oleic Acid (9c-18:1) (% (w/w))	0164	1.1100	0.04000				3	724.99	0	
724	Oleic Acid (9c-18:1) (% (w/w))	2161	1.3570	0.10800				3	724.99	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	2161	1.5124	0.11990	1.5796	0.05164	0.11775	4	726.99	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0226	1.5724	0.11410	1.5796	0.05164	0.11775	4	726.99	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0676	1.5985	0.06700	1.5796	0.05164	0.11775	4	726.99	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0164	1.6350	0.17000	1.5796	0.05164	0.11775	4	726.99	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/w))	2161	0.07805	0.00630	0.10526	0.01870	0.00758	4	728.99	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/w))	0226	0.10850	0.00900	0.10526	0.01870	0.00758	4	728.99	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/w))	0164	0.11500	0.01000	0.10526	0.01870	0.00758	4	728.99	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/w))	0676	0.11950	0.00500	0.10526	0.01870	0.00758	4	728.99	0	
730	Arachidic Acid (20:0) (% (w/w))	0226	0.01250	0.00080				2	730.99	0	
730	Arachidic Acid (20:0) (% (w/w))	2161	0.01940	0.00120				2	730.99	0	
732	Gondoic Acid (11c-20:1) (% (w/w))	0226	0.01240	0.00060				2	732.99	0	
732	Gondoic Acid (11c-20:1) (% (w/w))	2161	0.02355	0.00070				2	732.99	0	
738	Mead Acid (11c,14c,17c-20:3) (% (w/w))	2161	< 0.01					0	738.99	5	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c) (% (w/w))	0226	0.00260	0.00020				1	740.99	0	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c) (% (w/w))	2161	< 0.01					1	740.99	5	
742	Behenic Acid (22:0) (% (w/w))	0226	0.00880	0.00100				3	742.99	0	
742	Behenic Acid (22:0) (% (w/w))	0164	0.01000	0.00000				3	742.99	0	
742	Behenic Acid (22:0) (% (w/w))	2161	0.01140	0.00080				3	742.99	0	
744	Erucic Acid (13c-22:1) (% (w/w))	0226	0.00000	0.00000				0	744.99	4	
744	Erucic Acid (13c-22:1) (% (w/w))	2161	< 0.01					0	744.99	5	
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c) (% (w/w))	0226	0.00000	0.00000				0	746.99	4	
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c) (% (w/w))	2161	< 0.01					0	746.99	5	
748	Lignoceric Acid (24:0) (% (w/w))	0226	0.00775	0.00130				2	748.99	0	
748	Lignoceric Acid (24:0) (% (w/w))	2161	0.00960	0.00100				2	748.99	0	
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22c) (% (w/w))	0226	0.00000	0.00000				0	750.99	4	

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16	2161	< 0.01					0	750.99	5	
752	Nervonic Acid (24:1) isomers (% (w/w))	0226	0.00000	0.00000				0	752.99	4	
752	Nervonic Acid (24:1) isomers (% (w/w))	2161	< 0.01					0	752.99	5	
754	Total n-3 Polyunsaturated (Omega-3) Fatty Aci	0164	0.12500	0.01000				2	754.99	0	
754	Total n-3 Polyunsaturated (Omega-3) Fatty Aci	0676	0.12500	0.01000				2	754.99	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Aci	0676	1.6050	0.07000				2	756.99	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Aci	0164	1.6200	0.14000				2	756.99	0	
758	Total Saturated Fatty Acids (% (w/w))	0164	0.98000	0.08000				1	758.99	0	
762	Total Monounsaturated Fatty Acids (% (w/w))	0164	1.1500	0.06000				1	762.99	0	
766	Total Polyunsaturated Fatty Acids (% (w/w))	0164	1.7650	0.17000				1	766.99	0	
770	Total Fat (equivalent to NLEA) (% (w/w))	0164	4.1100	0.34000				1	770.99	0	
772	Total Fatty Acids (% (w/w))	0226	3.5785	0.26270				2	772.99	0	
772	Total Fatty Acids (% (w/w))	0164	3.9300	0.32000				2	772.99	0	

Notes: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2 and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = Rejected for duplicates too far apart, 2 = Rejected as outlier, 5 = A reporting limit and 4 = zeros submitted as values. Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). Flag 3 indicates not used in statistics.