



AAFCO
Proficiency Testing Program



Equine Crumbles

Animal Feed Scheme

Labs Reporting: 192

Test Material Code # 201890

Analyte Proficiency Testing Report

Issue Date : 11/30/2018

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
000	Urea (%)	0278	0.1000	0.0000				1		000.02	0
001	Loss on Drying (%)	0638	2.400	0.0000	7.266	0.3072	0.1378	78	-15.84	001.07	0
001	Loss on Drying (%)	0618	5.400	0.0800	7.266	0.3072	0.1378	78	-6.08	001.99	0
001	Loss on Drying (%)	2155	5.800	0.5400	7.266	0.3072	0.1378	78	-4.77	001.99	0
001	Loss on Drying (%)	0510	6.350	0.1000	7.266	0.3072	0.1378	78	-2.98	001.99	0
001	Loss on Drying (%)	0889	6.670	0.0200	7.266	0.3072	0.1378	78	-1.94	001.99	0
001	Loss on Drying (%)	0536	6.745	0.0900	7.266	0.3072	0.1378	78	-1.70	001.99	0
001	Loss on Drying (%)	0019	6.750	0.3000	7.266	0.3072	0.1378	78	-1.68	001.07	0
001	Loss on Drying (%)	0226	6.840	0.0000	7.266	0.3072	0.1378	78	-1.39	001.07	0
001	Loss on Drying (%)	0852	6.855	0.0100	7.266	0.3072	0.1378	78	-1.34	001.99	0
001	Loss on Drying (%)	0309	6.910	0.0200	7.266	0.3072	0.1378	78	-1.16	001.00	0
001	Loss on Drying (%)	0683	6.915	0.1100	7.266	0.3072	0.1378	78	-1.14	001.07	0
001	Loss on Drying (%)	0345	6.955	0.0300	7.266	0.3072	0.1378	78	-1.01	001.07	0
001	Loss on Drying (%)	2146	7.000	0.0200	7.266	0.3072	0.1378	78	-0.87	001.07	0
001	Loss on Drying (%)	0630	7.010	0.1400	7.266	0.3072	0.1378	78	-0.83	001.99	0
001	Loss on Drying (%)	0034	7.015	0.2100	7.266	0.3072	0.1378	78	-0.82	001.99	0
001	Loss on Drying (%)	0142	7.045	0.0500	7.266	0.3072	0.1378	78	-0.72	001.07	0
001	Loss on Drying (%)	0505	7.045	0.4900	7.266	0.3072	0.1378	78	-0.72	001.07	0
001	Loss on Drying (%)	2126	7.050	0.0200	7.266	0.3072	0.1378	78	-0.70	001.99	0
001	Loss on Drying (%)	0610	7.055	0.0100	7.266	0.3072	0.1378	78	-0.69	001.05	0
001	Loss on Drying (%)	0918	7.075	0.1500	7.266	0.3072	0.1378	78	-0.62	001.99	0
001	Loss on Drying (%)	0187	7.080	0.0400	7.266	0.3072	0.1378	78	-0.61	001.07	0
001	Loss on Drying (%)	0619	7.100	0.0000	7.266	0.3072	0.1378	78	-0.54	001.03	0
001	Loss on Drying (%)	0609	7.100	0.2000	7.266	0.3072	0.1378	78	-0.54	001.07	0
001	Loss on Drying (%)	2145	7.105	0.0100	7.266	0.3072	0.1378	78	-0.53	001.07	0
001	Loss on Drying (%)	0089	7.115	0.0100	7.266	0.3072	0.1378	78	-0.49	001.07	0
001	Loss on Drying (%)	0353	7.120	0.1800	7.266	0.3072	0.1378	78	-0.48	001.07	0
001	Loss on Drying (%)	0148	7.145	0.0100	7.266	0.3072	0.1378	78	-0.39	001.07	0
001	Loss on Drying (%)	2012	7.160	0.0400	7.266	0.3072	0.1378	78	-0.35	001.07	0
001	Loss on Drying (%)	0027	7.166	0.0830	7.266	0.3072	0.1378	78	-0.33	001.00	0
001	Loss on Drying (%)	0644	7.170	0.2000	7.266	0.3072	0.1378	78	-0.31	001.05	0
001	Loss on Drying (%)	0675	7.170	0.0800	7.266	0.3072	0.1378	78	-0.31	001.07	0
001	Loss on Drying (%)	2059	7.190	0.0400	7.266	0.3072	0.1378	78	-0.25	001.07	0
001	Loss on Drying (%)	0016	7.190	0.0000	7.266	0.3072	0.1378	78	-0.25	001.00	0
001	Loss on Drying (%)	0948	7.200	0.1000	7.266	0.3072	0.1378	78	-0.22	001.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
001	Loss on Drying (%)	0656	7.220	0.5400	7.266	0.3072	0.1378	78	-0.15	001.99	0
001	Loss on Drying (%)	0003	7.240	0.0800	7.266	0.3072	0.1378	78	-0.09	001.07	0
001	Loss on Drying (%)	0083	7.240	0.0600	7.266	0.3072	0.1378	78	-0.09	001.07	0
001	Loss on Drying (%)	2009	7.248	0.0065	7.266	0.3072	0.1378	78	-0.06	001.07	0
001	Loss on Drying (%)	0186	7.260	0.1000	7.266	0.3072	0.1378	78	-0.02	001.99	0
001	Loss on Drying (%)	0504	7.275	0.0300	7.266	0.3072	0.1378	78	0.03	001.00	0
001	Loss on Drying (%)	0098	7.285	0.2300	7.266	0.3072	0.1378	78	0.06	001.07	0
001	Loss on Drying (%)	2190	7.295	0.0100	7.266	0.3072	0.1378	78	0.09	001.07	0
001	Loss on Drying (%)	2195	7.300	0.0200	7.266	0.3072	0.1378	78	0.11	001.07	0
001	Loss on Drying (%)	2062	7.311	0.0140	7.266	0.3072	0.1378	78	0.15	001.03	0
001	Loss on Drying (%)	0171	7.315	0.0100	7.266	0.3072	0.1378	78	0.16	001.07	0
001	Loss on Drying (%)	0612	7.350	0.7200	7.266	0.3072	0.1378	78	0.27	001.99	0
001	Loss on Drying (%)	0868	7.355	0.0100	7.266	0.3072	0.1378	78	0.29	001.03	0
001	Loss on Drying (%)	0407	7.356	0.0458	7.266	0.3072	0.1378	78	0.29	001.07	0
001	Loss on Drying (%)	0049	7.365	0.0100	7.266	0.3072	0.1378	78	0.32	001.07	0
001	Loss on Drying (%)	0066	7.375	0.1700	7.266	0.3072	0.1378	78	0.35	001.07	0
001	Loss on Drying (%)	0849	7.395	0.1900	7.266	0.3072	0.1378	78	0.42	001.07	0
001	Loss on Drying (%)	0689	7.400	0.0000	7.266	0.3072	0.1378	78	0.44	001.07	0
001	Loss on Drying (%)	0910	7.400	0.2000	7.266	0.3072	0.1378	78	0.44	001.99	0
001	Loss on Drying (%)	0571	7.410	0.0600	7.266	0.3072	0.1378	78	0.47	001.07	0
001	Loss on Drying (%)	0028	7.430	0.0600	7.266	0.3072	0.1378	78	0.53	001.07	0
001	Loss on Drying (%)	0720	7.430	0.3400	7.266	0.3072	0.1378	78	0.53	001.99	0
001	Loss on Drying (%)	0590	7.445	0.0500	7.266	0.3072	0.1378	78	0.58	001.07	0
001	Loss on Drying (%)	0037	7.450	0.0200	7.266	0.3072	0.1378	78	0.60	001.99	0
001	Loss on Drying (%)	0693	7.457	0.3690	7.266	0.3072	0.1378	78	0.62	001.07	0
001	Loss on Drying (%)	0876	7.463	0.0596	7.266	0.3072	0.1378	78	0.64	001.07	0
001	Loss on Drying (%)	2053	7.475	0.0700	7.266	0.3072	0.1378	78	0.68	001.07	0
001	Loss on Drying (%)	0878	7.480	0.0400	7.266	0.3072	0.1378	78	0.70	001.03	0
001	Loss on Drying (%)	0682	7.480	0.0000	7.266	0.3072	0.1378	78	0.70	001.07	0
001	Loss on Drying (%)	0912	7.485	0.2700	7.266	0.3072	0.1378	78	0.71	001.99	0
001	Loss on Drying (%)	0278	7.585	0.0900	7.266	0.3072	0.1378	78	1.04	001.07	0
001	Loss on Drying (%)	2144	7.605	0.0300	7.266	0.3072	0.1378	78	1.10	001.07	0
001	Loss on Drying (%)	0872	7.610	0.0400	7.266	0.3072	0.1378	78	1.12	001.07	0
001	Loss on Drying (%)	2109	7.620	0.1800	7.266	0.3072	0.1378	78	1.15	001.07	0
001	Loss on Drying (%)	0843	7.635	0.2100	7.266	0.3072	0.1378	78	1.20	001.07	0
001	Loss on Drying (%)	2192	7.635	0.7100	7.266	0.3072	0.1378	78	1.20	001.07	0
001	Loss on Drying (%)	0074	7.675	0.4900	7.266	0.3072	0.1378	78	1.33	001.07	0
001	Loss on Drying (%)	0676	7.675	0.0100	7.266	0.3072	0.1378	78	1.33	001.99	0
001	Loss on Drying (%)	0357	7.687	0.1814	7.266	0.3072	0.1378	78	1.37	001.99	0
001	Loss on Drying (%)	0629	7.740	0.3000	7.266	0.3072	0.1378	78	1.54	001.99	0
001	Loss on Drying (%)	0169	7.770	0.0200	7.266	0.3072	0.1378	78	1.64	001.00	0
001	Loss on Drying (%)	0695	7.845	0.1700	7.266	0.3072	0.1378	78	1.88	001.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
001	Loss on Drying (%)	0019	7.965	0.2500	7.266	0.3072	0.1378	78	2.27	001.99	0
001	Loss on Drying (%)	0045	8.040	0.9000	7.266	0.3072	0.1378	78	2.52	001.07	0
001	Loss on Drying (%)	2081	7.025	1.590	7.266	0.3072	0.1378	78	-0.79	001.99	1
001	Loss on Drying (%)	0560	7.150	1.920	7.266	0.3072	0.1378	78	-0.38	001.99	1
001	Loss on Drying (%)	0893	14.15	0.0000	7.266	0.3072	0.1378	78	22.41	001.03	2
002	Protein, Crude (%)	0638	7.550	0.0200	14.18	0.3046	0.1178	184	-21.76	002.04	0
002	Protein, Crude (%)	0893	7.610	0.0000	14.18	0.3046	0.1178	184	-21.56	002.05	0
002	Protein, Crude (%)	0852	12.38	0.2200	14.18	0.3046	0.1178	184	-5.90	002.11	0
002	Protein, Crude (%)	0889	12.63	0.7600	14.18	0.3046	0.1178	184	-5.08	002.11	0
002	Protein, Crude (%)	0723	13.53	0.0000	14.18	0.3046	0.1178	184	-2.13	002.01	0
002	Protein, Crude (%)	2155	13.55	0.3780	14.18	0.3046	0.1178	184	-2.07	002.06	0
002	Protein, Crude (%)	0685	13.55	0.0600	14.18	0.3046	0.1178	184	-2.06	002.01	0
002	Protein, Crude (%)	0910	13.60	0.4000	14.18	0.3046	0.1178	184	-1.90	002.01	0
002	Protein, Crude (%)	2190	13.63	0.1000	14.18	0.3046	0.1178	184	-1.80	002.05	0
002	Protein, Crude (%)	2145	13.67	0.1600	14.18	0.3046	0.1178	184	-1.67	002.05	0
002	Protein, Crude (%)	0874	13.70	0.2300	14.18	0.3046	0.1178	184	-1.59	002.04	0
002	Protein, Crude (%)	0720	13.70	0.0000	14.18	0.3046	0.1178	184	-1.57	002.06	0
002	Protein, Crude (%)	2190	13.70	0.1200	14.18	0.3046	0.1178	184	-1.57	002.06	0
002	Protein, Crude (%)	0036	13.74	0.0984	14.18	0.3046	0.1178	184	-1.43	002.02	0
002	Protein, Crude (%)	2073	13.76	0.0600	14.18	0.3046	0.1178	184	-1.37	002.05	0
002	Protein, Crude (%)	0914	13.77	0.1900	14.18	0.3046	0.1178	184	-1.36	002.01	0
002	Protein, Crude (%)	2223	13.77	0.0000	14.18	0.3046	0.1178	184	-1.34	002.01	0
002	Protein, Crude (%)	0536	13.78	0.4100	14.18	0.3046	0.1178	184	-1.33	002.05	0
002	Protein, Crude (%)	0169	13.78	0.1000	14.18	0.3046	0.1178	184	-1.31	002.02	0
002	Protein, Crude (%)	0885	13.80	0.0007	14.18	0.3046	0.1178	184	-1.25	002.05	0
002	Protein, Crude (%)	0870	13.80	0.1220	14.18	0.3046	0.1178	184	-1.23	002.01	0
002	Protein, Crude (%)	0656	13.82	0.0600	14.18	0.3046	0.1178	184	-1.18	002.01	0
002	Protein, Crude (%)	0504	13.83	0.0800	14.18	0.3046	0.1178	184	-1.14	002.04	0
002	Protein, Crude (%)	0939	13.83	0.1200	14.18	0.3046	0.1178	184	-1.14	002.05	0
002	Protein, Crude (%)	0541	13.83	0.1800	14.18	0.3046	0.1178	184	-1.14	002.06	0
002	Protein, Crude (%)	0683	13.85	0.0200	14.18	0.3046	0.1178	184	-1.08	002.05	0
002	Protein, Crude (%)	0619	13.85	0.1000	14.18	0.3046	0.1178	184	-1.08	002.05	0
002	Protein, Crude (%)	0955	13.85	0.1000	14.18	0.3046	0.1178	184	-1.08	002.05	0
002	Protein, Crude (%)	0354	13.86	0.0300	14.18	0.3046	0.1178	184	-1.06	002.05	0
002	Protein, Crude (%)	0693	13.86	0.0700	14.18	0.3046	0.1178	184	-1.05	002.06	0
002	Protein, Crude (%)	2192	13.87	0.4100	14.18	0.3046	0.1178	184	-1.03	002.05	0
002	Protein, Crude (%)	0098	13.88	0.0200	14.18	0.3046	0.1178	184	-0.98	002.08	0
002	Protein, Crude (%)	0018	13.89	0.0300	14.18	0.3046	0.1178	184	-0.96	002.06	0
002	Protein, Crude (%)	2009	13.89	0.0271	14.18	0.3046	0.1178	184	-0.96	002.05	0
002	Protein, Crude (%)	0653	13.90	0.0800	14.18	0.3046	0.1178	184	-0.91	002.01	0
002	Protein, Crude (%)	0676	13.91	0.1900	14.18	0.3046	0.1178	184	-0.90	002.06	0
002	Protein, Crude (%)	0939	13.91	0.1500	14.18	0.3046	0.1178	184	-0.90	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein, Crude (%)	2109	13.92	0.1400	14.18	0.3046	0.1178	184	-0.85	002.05	0
002	Protein, Crude (%)	0042	13.93	0.1900	14.18	0.3046	0.1178	184	-0.83	002.02	0
002	Protein, Crude (%)	0539	13.93	0.2200	14.18	0.3046	0.1178	184	-0.82	002.06	0
002	Protein, Crude (%)	0682	13.93	0.0000	14.18	0.3046	0.1178	184	-0.82	002.06	0
002	Protein, Crude (%)	0964	13.93	0.1760	14.18	0.3046	0.1178	184	-0.81	002.06	0
002	Protein, Crude (%)	0968	13.94	0.0150	14.18	0.3046	0.1178	184	-0.78	002.01	0
002	Protein, Crude (%)	0689	13.95	0.1000	14.18	0.3046	0.1178	184	-0.75	002.05	0
002	Protein, Crude (%)	2022	13.96	0.0300	14.18	0.3046	0.1178	184	-0.73	002.05	0
002	Protein, Crude (%)	0049	13.96	0.3600	14.18	0.3046	0.1178	184	-0.72	002.06	0
002	Protein, Crude (%)	0345	13.97	0.0670	14.18	0.3046	0.1178	184	-0.70	002.06	0
002	Protein, Crude (%)	0848	13.98	0.0100	14.18	0.3046	0.1178	184	-0.67	002.06	0
002	Protein, Crude (%)	0598	13.99	0.1300	14.18	0.3046	0.1178	184	-0.64	002.06	0
002	Protein, Crude (%)	0164	14.00	0.0000	14.18	0.3046	0.1178	184	-0.59	002.01	0
002	Protein, Crude (%)	0675	14.01	0.0200	14.18	0.3046	0.1178	184	-0.55	002.05	0
002	Protein, Crude (%)	0881	14.01	0.0300	14.18	0.3046	0.1178	184	-0.55	002.05	0
002	Protein, Crude (%)	0100	14.03	0.1000	14.18	0.3046	0.1178	184	-0.49	002.06	0
002	Protein, Crude (%)	0066	14.04	0.1300	14.18	0.3046	0.1178	184	-0.47	002.02	0
002	Protein, Crude (%)	0695	14.04	0.2190	14.18	0.3046	0.1178	184	-0.47	002.06	0
002	Protein, Crude (%)	0171	14.05	0.0500	14.18	0.3046	0.1178	184	-0.44	002.06	0
002	Protein, Crude (%)	0123	14.05	0.1000	14.18	0.3046	0.1178	184	-0.42	002.06	0
002	Protein, Crude (%)	0144	14.05	0.0400	14.18	0.3046	0.1178	184	-0.42	002.06	0
002	Protein, Crude (%)	0089	14.06	0.0100	14.18	0.3046	0.1178	184	-0.41	002.06	0
002	Protein, Crude (%)	2006	14.06	0.0600	14.18	0.3046	0.1178	184	-0.39	002.05	0
002	Protein, Crude (%)	0853	14.07	0.3900	14.18	0.3046	0.1178	184	-0.37	002.06	0
002	Protein, Crude (%)	0610	14.08	0.2500	14.18	0.3046	0.1178	184	-0.34	002.01	0
002	Protein, Crude (%)	2188	14.09	0.0500	14.18	0.3046	0.1178	184	-0.31	002.01	0
002	Protein, Crude (%)	0878	14.09	0.0100	14.18	0.3046	0.1178	184	-0.31	002.05	0
002	Protein, Crude (%)	0202	14.09	0.0100	14.18	0.3046	0.1178	184	-0.29	002.06	0
002	Protein, Crude (%)	0868	14.10	0.0300	14.18	0.3046	0.1178	184	-0.27	002.05	0
002	Protein, Crude (%)	0205	14.10	0.1800	14.18	0.3046	0.1178	184	-0.26	002.06	0
002	Protein, Crude (%)	0955	14.10	0.4000	14.18	0.3046	0.1178	184	-0.26	002.06	0
002	Protein, Crude (%)	0897	14.11	0.0000	14.18	0.3046	0.1178	184	-0.23	002.05	0
002	Protein, Crude (%)	0629	14.11	0.0200	14.18	0.3046	0.1178	184	-0.23	002.10	0
002	Protein, Crude (%)	0849	14.12	0.1920	14.18	0.3046	0.1178	184	-0.21	002.05	0
002	Protein, Crude (%)	0309	14.12	0.1050	14.18	0.3046	0.1178	184	-0.21	002.06	0
002	Protein, Crude (%)	0034	14.12	0.6520	14.18	0.3046	0.1178	184	-0.20	002.06	0
002	Protein, Crude (%)	2196	14.12	0.0000	14.18	0.3046	0.1178	184	-0.19	002.01	0
002	Protein, Crude (%)	2181	14.12	0.0600	14.18	0.3046	0.1178	184	-0.19	002.05	0
002	Protein, Crude (%)	0968	14.12	0.0520	14.18	0.3046	0.1178	184	-0.19	002.06	0
002	Protein, Crude (%)	0164	14.13	0.1300	14.18	0.3046	0.1178	184	-0.18	002.06	0
002	Protein, Crude (%)	0610	14.13	0.0500	14.18	0.3046	0.1178	184	-0.18	002.06	0
002	Protein, Crude (%)	2129	14.14	0.0500	14.18	0.3046	0.1178	184	-0.14	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein, Crude (%)	2193	14.14	0.1400	14.18	0.3046	0.1178	184	-0.13	002.01	0
002	Protein, Crude (%)	0505	14.15	0.0900	14.18	0.3046	0.1178	184	-0.11	002.06	0
002	Protein, Crude (%)	2199	14.15	0.1000	14.18	0.3046	0.1178	184	-0.09	002.05	0
002	Protein, Crude (%)	0010	14.15	0.1000	14.18	0.3046	0.1178	184	-0.09	002.06	0
002	Protein, Crude (%)	0047	14.15	0.1000	14.18	0.3046	0.1178	184	-0.09	002.06	0
002	Protein, Crude (%)	0278	14.15	0.1000	14.18	0.3046	0.1178	184	-0.09	002.06	0
002	Protein, Crude (%)	0956	14.15	0.3000	14.18	0.3046	0.1178	184	-0.09	002.06	0
002	Protein, Crude (%)	0880	14.15	0.1510	14.18	0.3046	0.1178	184	-0.08	002.06	0
002	Protein, Crude (%)	0187	14.16	0.0700	14.18	0.3046	0.1178	184	-0.08	002.04	0
002	Protein, Crude (%)	0074	14.16	0.0500	14.18	0.3046	0.1178	184	-0.08	002.06	0
002	Protein, Crude (%)	0948	14.17	0.0300	14.18	0.3046	0.1178	184	-0.04	002.05	0
002	Protein, Crude (%)	2150	14.17	0.0100	14.18	0.3046	0.1178	184	-0.04	002.06	0
002	Protein, Crude (%)	0646	14.17	0.3200	14.18	0.3046	0.1178	184	-0.03	002.06	0
002	Protein, Crude (%)	0674	14.18	0.0000	14.18	0.3046	0.1178	184	0.00	002.05	0
002	Protein, Crude (%)	0142	14.18	0.0200	14.18	0.3046	0.1178	184	0.00	002.06	0
002	Protein, Crude (%)	0889	14.19	0.1300	14.18	0.3046	0.1178	184	0.02	002.01	0
002	Protein, Crude (%)	0042	14.19	0.3500	14.18	0.3046	0.1178	184	0.02	002.06	0
002	Protein, Crude (%)	2215	14.19	0.0900	14.18	0.3046	0.1178	184	0.02	002.06	0
002	Protein, Crude (%)	0626	14.19	0.0010	14.18	0.3046	0.1178	184	0.04	002.06	0
002	Protein, Crude (%)	0003	14.20	0.0500	14.18	0.3046	0.1178	184	0.05	002.06	0
002	Protein, Crude (%)	0868	14.20	0.0100	14.18	0.3046	0.1178	184	0.05	002.06	0
002	Protein, Crude (%)	0175	14.20	0.0000	14.18	0.3046	0.1178	184	0.07	002.06	0
002	Protein, Crude (%)	0884	14.20	0.0000	14.18	0.3046	0.1178	184	0.07	002.06	0
002	Protein, Crude (%)	0644	14.21	0.1500	14.18	0.3046	0.1178	184	0.09	002.06	0
002	Protein, Crude (%)	0152	14.22	0.0100	14.18	0.3046	0.1178	184	0.12	002.01	0
002	Protein, Crude (%)	0687	14.22	0.1700	14.18	0.3046	0.1178	184	0.12	002.06	0
002	Protein, Crude (%)	0226	14.22	0.0800	14.18	0.3046	0.1178	184	0.14	002.06	0
002	Protein, Crude (%)	2198	14.23	0.0600	14.18	0.3046	0.1178	184	0.17	002.01	0
002	Protein, Crude (%)	0948	14.23	0.0400	14.18	0.3046	0.1178	184	0.17	002.06	0
002	Protein, Crude (%)	0870	14.24	0.0563	14.18	0.3046	0.1178	184	0.19	002.06	0
002	Protein, Crude (%)	0859	14.24	0.0440	14.18	0.3046	0.1178	184	0.20	002.06	0
002	Protein, Crude (%)	0592	14.25	0.1500	14.18	0.3046	0.1178	184	0.22	002.06	0
002	Protein, Crude (%)	0006	14.25	0.0200	14.18	0.3046	0.1178	184	0.23	002.06	0
002	Protein, Crude (%)	0016	14.25	0.1000	14.18	0.3046	0.1178	184	0.23	002.06	0
002	Protein, Crude (%)	0098	14.25	0.1000	14.18	0.3046	0.1178	184	0.23	002.06	0
002	Protein, Crude (%)	0619	14.25	0.1000	14.18	0.3046	0.1178	184	0.23	002.06	0
002	Protein, Crude (%)	0692	14.25	0.7000	14.18	0.3046	0.1178	184	0.23	002.06	0
002	Protein, Crude (%)	0300	14.27	0.0400	14.18	0.3046	0.1178	184	0.30	002.06	0
002	Protein, Crude (%)	2081	14.28	0.1100	14.18	0.3046	0.1178	184	0.32	002.06	0
002	Protein, Crude (%)	0028	14.29	0.0700	14.18	0.3046	0.1178	184	0.35	002.01	0
002	Protein, Crude (%)	0353	14.30	0.0400	14.18	0.3046	0.1178	184	0.40	002.06	0
002	Protein, Crude (%)	0504	14.30	0.2800	14.18	0.3046	0.1178	184	0.40	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0952	14.30	0.1070	14.18	0.3046	0.1178	184	0.41	002.05	0
002	Protein, Crude (%)	0857	14.31	0.1100	14.18	0.3046	0.1178	184	0.41	002.06	0
002	Protein, Crude (%)	0263	14.31	0.0293	14.18	0.3046	0.1178	184	0.42	002.06	0
002	Protein, Crude (%)	0675	14.31	0.0400	14.18	0.3046	0.1178	184	0.43	002.06	0
002	Protein, Crude (%)	0357	14.31	0.0000	14.18	0.3046	0.1178	184	0.44	002.06	0
002	Protein, Crude (%)	2126	14.32	0.0600	14.18	0.3046	0.1178	184	0.46	002.06	0
002	Protein, Crude (%)	0027	14.32	0.3110	14.18	0.3046	0.1178	184	0.47	002.06	0
002	Protein, Crude (%)	0358	14.35	0.1800	14.18	0.3046	0.1178	184	0.56	002.06	0
002	Protein, Crude (%)	0265	14.35	0.1000	14.18	0.3046	0.1178	184	0.56	002.06	0
002	Protein, Crude (%)	0590	14.35	0.1000	14.18	0.3046	0.1178	184	0.56	002.06	0
002	Protein, Crude (%)	0571	14.35	0.0250	14.18	0.3046	0.1178	184	0.56	002.06	0
002	Protein, Crude (%)	0354	14.36	0.0200	14.18	0.3046	0.1178	184	0.60	002.06	0
002	Protein, Crude (%)	0589	14.36	0.0000	14.18	0.3046	0.1178	184	0.60	002.06	0
002	Protein, Crude (%)	2075	14.36	0.0800	14.18	0.3046	0.1178	184	0.60	002.06	0
002	Protein, Crude (%)	0108	14.37	0.1100	14.18	0.3046	0.1178	184	0.61	002.06	0
002	Protein, Crude (%)	0148	14.37	0.5300	14.18	0.3046	0.1178	184	0.61	002.06	0
002	Protein, Crude (%)	0872	14.37	0.0300	14.18	0.3046	0.1178	184	0.61	002.06	0
002	Protein, Crude (%)	2053	14.37	0.0900	14.18	0.3046	0.1178	184	0.61	002.06	0
002	Protein, Crude (%)	0035	14.40	0.0100	14.18	0.3046	0.1178	184	0.71	002.06	0
002	Protein, Crude (%)	2200	14.40	0.0000	14.18	0.3046	0.1178	184	0.73	002.05	0
002	Protein, Crude (%)	0407	14.41	0.0680	14.18	0.3046	0.1178	184	0.75	002.06	0
002	Protein, Crude (%)	2195	14.41	0.1400	14.18	0.3046	0.1178	184	0.76	002.01	0
002	Protein, Crude (%)	2022	14.41	0.0760	14.18	0.3046	0.1178	184	0.77	002.06	0
002	Protein, Crude (%)	0553	14.42	0.2700	14.18	0.3046	0.1178	184	0.78	002.11	0
002	Protein, Crude (%)	0004	14.42	0.1000	14.18	0.3046	0.1178	184	0.79	002.06	0
002	Protein, Crude (%)	0037	14.42	0.2800	14.18	0.3046	0.1178	184	0.79	002.06	0
002	Protein, Crude (%)	2089	14.43	0.2100	14.18	0.3046	0.1178	184	0.81	002.06	0
002	Protein, Crude (%)	0726	14.43	0.0000	14.18	0.3046	0.1178	184	0.83	002.06	0
002	Protein, Crude (%)	0009	14.45	0.0100	14.18	0.3046	0.1178	184	0.87	002.06	0
002	Protein, Crude (%)	0574	14.45	0.1100	14.18	0.3046	0.1178	184	0.87	002.06	0
002	Protein, Crude (%)	0019	14.49	0.3800	14.18	0.3046	0.1178	184	1.02	002.06	0
002	Protein, Crude (%)	0083	14.49	0.0400	14.18	0.3046	0.1178	184	1.02	002.06	0
002	Protein, Crude (%)	0036	14.50	0.1500	14.18	0.3046	0.1178	184	1.04	002.06	0
002	Protein, Crude (%)	0596	14.50	0.0100	14.18	0.3046	0.1178	184	1.04	002.06	0
002	Protein, Crude (%)	0033	14.50	0.0000	14.18	0.3046	0.1178	184	1.05	002.06	0
002	Protein, Crude (%)	0045	14.50	0.2000	14.18	0.3046	0.1178	184	1.05	002.06	0
002	Protein, Crude (%)	0510	14.50	0.0000	14.18	0.3046	0.1178	184	1.05	002.06	0
002	Protein, Crude (%)	0910	14.50	0.0000	14.18	0.3046	0.1178	184	1.05	002.06	0
002	Protein, Crude (%)	2146	14.52	0.1500	14.18	0.3046	0.1178	184	1.10	002.05	0
002	Protein, Crude (%)	0609	14.52	0.1800	14.18	0.3046	0.1178	184	1.12	002.06	0
002	Protein, Crude (%)	0630	14.53	0.0200	14.18	0.3046	0.1178	184	1.15	002.06	0
002	Protein, Crude (%)	2059	14.55	0.0656	14.18	0.3046	0.1178	184	1.22	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein, Crude (%)	0190	14.56	0.2300	14.18	0.3046	0.1178	184	1.24	002.06	0
002	Protein, Crude (%)	0843	14.56	0.0900	14.18	0.3046	0.1178	184	1.24	002.06	0
002	Protein, Crude (%)	0425	14.56	0.0200	14.18	0.3046	0.1178	184	1.25	002.06	0
002	Protein, Crude (%)	0941	14.57	0.1600	14.18	0.3046	0.1178	184	1.28	002.06	0
002	Protein, Crude (%)	0511	14.59	0.1400	14.18	0.3046	0.1178	184	1.35	002.06	0
002	Protein, Crude (%)	0186	14.60	0.2000	14.18	0.3046	0.1178	184	1.38	002.06	0
002	Protein, Crude (%)	0861	14.60	0.1400	14.18	0.3046	0.1178	184	1.38	002.06	0
002	Protein, Crude (%)	2076	14.61	0.0660	14.18	0.3046	0.1178	184	1.43	002.06	0
002	Protein, Crude (%)	0918	14.62	0.0200	14.18	0.3046	0.1178	184	1.45	002.06	0
002	Protein, Crude (%)	0242	14.64	0.0200	14.18	0.3046	0.1178	184	1.51	002.06	0
002	Protein, Crude (%)	0001	14.66	0.1400	14.18	0.3046	0.1178	184	1.58	002.06	0
002	Protein, Crude (%)	0229	14.67	0.0500	14.18	0.3046	0.1178	184	1.60	002.06	0
002	Protein, Crude (%)	0618	14.68	0.0180	14.18	0.3046	0.1178	184	1.66	002.06	0
002	Protein, Crude (%)	0520	14.75	0.0300	14.18	0.3046	0.1178	184	1.86	002.06	0
002	Protein, Crude (%)	0876	14.76	0.1321	14.18	0.3046	0.1178	184	1.91	002.06	0
002	Protein, Crude (%)	0119	15.18	0.0000	14.18	0.3046	0.1178	184	3.29	002.06	0
002	Protein, Crude (%)	0953	15.22	0.5800	14.18	0.3046	0.1178	184	3.42	002.06	0
002	Protein, Crude (%)	0852	15.24	0.1700	14.18	0.3046	0.1178	184	3.47	002.05	0
002	Protein, Crude (%)	0612	16.01	0.5100	14.18	0.3046	0.1178	184	6.00	002.06	0
002	Protein, Crude (%)	0405	16.58	0.0000	14.18	0.3046	0.1178	184	7.88	002.04	0
002	Protein, Crude (%)	0720	19.80	0.1100	14.18	0.3046	0.1178	184	18.44	002.11	0
002	Protein, Crude (%)	2128	12.60	2.887	14.18	0.3046	0.1178	184	-5.19	002.03	1
002	Protein, Crude (%)	0643	0.1432	0.0008	14.18	0.3046	0.1178	184	-46.08	002.99	2
003	Fat, Crude (%)	0880	0.9000	1.400	1.788	0.2928	0.1156	117	-3.03	003.99	0
003	Fat, Crude (%)	0952	1.080	0.3200	1.788	0.2928	0.1156	117	-2.42	003.10	0
003	Fat, Crude (%)	2129	1.130	0.0200	1.788	0.2928	0.1156	117	-2.25	003.14	0
003	Fat, Crude (%)	0720	1.180	0.0200	1.788	0.2928	0.1156	117	-2.08	003.11	0
003	Fat, Crude (%)	0175	1.185	0.0500	1.788	0.2928	0.1156	117	-2.06	003.14	0
003	Fat, Crude (%)	0242	1.250	0.0000	1.788	0.2928	0.1156	117	-1.84	003.10	0
003	Fat, Crude (%)	0345	1.325	0.1300	1.788	0.2928	0.1156	117	-1.58	003.00	0
003	Fat, Crude (%)	0861	1.345	0.0300	1.788	0.2928	0.1156	117	-1.51	003.10	0
003	Fat, Crude (%)	2073	1.350	0.0000	1.788	0.2928	0.1156	117	-1.49	003.10	0
003	Fat, Crude (%)	0263	1.359	0.0090	1.788	0.2928	0.1156	117	-1.46	003.06	0
003	Fat, Crude (%)	0893	1.360	0.0000	1.788	0.2928	0.1156	117	-1.46	003.12	0
003	Fat, Crude (%)	0098	1.380	0.1000	1.788	0.2928	0.1156	117	-1.39	003.10	0
003	Fat, Crude (%)	0843	1.410	0.2000	1.788	0.2928	0.1156	117	-1.29	003.14	0
003	Fat, Crude (%)	0553	1.435	0.1900	1.788	0.2928	0.1156	117	-1.20	003.10	0
003	Fat, Crude (%)	0619	1.445	0.1700	1.788	0.2928	0.1156	117	-1.17	003.10	0
003	Fat, Crude (%)	0405	1.470	0.0000	1.788	0.2928	0.1156	117	-1.09	003.11	0
003	Fat, Crude (%)	0897	1.485	0.0300	1.788	0.2928	0.1156	117	-1.03	003.10	0
003	Fat, Crude (%)	2193	1.505	0.0100	1.788	0.2928	0.1156	117	-0.97	003.10	0
003	Fat, Crude (%)	0144	1.520	0.0000	1.788	0.2928	0.1156	117	-0.91	003.14	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat, Crude (%)	2190	1.530	0.0200	1.788	0.2928	0.1156	117	-0.88	003.00	0
003	Fat, Crude (%)	0520	1.540	0.3600	1.788	0.2928	0.1156	117	-0.85	003.14	0
003	Fat, Crude (%)	0186	1.560	0.1800	1.788	0.2928	0.1156	117	-0.78	003.14	0
003	Fat, Crude (%)	0100	1.585	0.2500	1.788	0.2928	0.1156	117	-0.69	003.10	0
003	Fat, Crude (%)	0358	1.585	0.0900	1.788	0.2928	0.1156	117	-0.69	003.14	0
003	Fat, Crude (%)	2076	1.593	0.0299	1.788	0.2928	0.1156	117	-0.66	003.14	0
003	Fat, Crude (%)	0357	1.600	0.0000	1.788	0.2928	0.1156	117	-0.64	003.14	0
003	Fat, Crude (%)	0119	1.605	0.0700	1.788	0.2928	0.1156	117	-0.62	003.10	0
003	Fat, Crude (%)	0205	1.615	0.0430	1.788	0.2928	0.1156	117	-0.59	003.13	0
003	Fat, Crude (%)	0089	1.620	0.0200	1.788	0.2928	0.1156	117	-0.57	003.14	0
003	Fat, Crude (%)	0693	1.632	0.3970	1.788	0.2928	0.1156	117	-0.53	003.10	0
003	Fat, Crude (%)	0045	1.635	0.3500	1.788	0.2928	0.1156	117	-0.52	003.10	0
003	Fat, Crude (%)	0541	1.635	0.3100	1.788	0.2928	0.1156	117	-0.52	003.10	0
003	Fat, Crude (%)	2188	1.635	0.1900	1.788	0.2928	0.1156	117	-0.52	003.10	0
003	Fat, Crude (%)	0574	1.640	0.0600	1.788	0.2928	0.1156	117	-0.50	003.14	0
003	Fat, Crude (%)	0629	1.650	0.1000	1.788	0.2928	0.1156	117	-0.47	003.10	0
003	Fat, Crude (%)	0098	1.650	0.2400	1.788	0.2928	0.1156	117	-0.47	003.13	0
003	Fat, Crude (%)	0956	1.650	0.1000	1.788	0.2928	0.1156	117	-0.47	003.14	0
003	Fat, Crude (%)	0695	1.665	0.0300	1.788	0.2928	0.1156	117	-0.42	003.10	0
003	Fat, Crude (%)	0598	1.665	0.2100	1.788	0.2928	0.1156	117	-0.42	003.14	0
003	Fat, Crude (%)	0202	1.668	0.0000	1.788	0.2928	0.1156	117	-0.41	003.14	0
003	Fat, Crude (%)	0889	1.675	0.2100	1.788	0.2928	0.1156	117	-0.38	003.11	0
003	Fat, Crude (%)	0878	1.680	0.0600	1.788	0.2928	0.1156	117	-0.37	003.14	0
003	Fat, Crude (%)	2181	1.685	0.5100	1.788	0.2928	0.1156	117	-0.35	003.06	0
003	Fat, Crude (%)	0171	1.690	0.0800	1.788	0.2928	0.1156	117	-0.33	003.12	0
003	Fat, Crude (%)	0083	1.690	0.0600	1.788	0.2928	0.1156	117	-0.33	003.14	0
003	Fat, Crude (%)	0019	1.695	0.4300	1.788	0.2928	0.1156	117	-0.32	003.14	0
003	Fat, Crude (%)	0948	1.695	0.0100	1.788	0.2928	0.1156	117	-0.32	003.14	0
003	Fat, Crude (%)	0504	1.705	0.3100	1.788	0.2928	0.1156	117	-0.28	003.01	0
003	Fat, Crude (%)	0870	1.715	0.0527	1.788	0.2928	0.1156	117	-0.25	003.10	0
003	Fat, Crude (%)	0948	1.715	0.0700	1.788	0.2928	0.1156	117	-0.25	003.09	0
003	Fat, Crude (%)	2089	1.725	0.2100	1.788	0.2928	0.1156	117	-0.21	003.13	0
003	Fat, Crude (%)	2022	1.740	0.0600	1.788	0.2928	0.1156	117	-0.16	003.10	0
003	Fat, Crude (%)	0868	1.755	0.0300	1.788	0.2928	0.1156	117	-0.11	003.10	0
003	Fat, Crude (%)	0265	1.760	0.1400	1.788	0.2928	0.1156	117	-0.09	003.14	0
003	Fat, Crude (%)	0407	1.761	0.1155	1.788	0.2928	0.1156	117	-0.09	003.14	0
003	Fat, Crude (%)	0425	1.765	0.0900	1.788	0.2928	0.1156	117	-0.08	003.06	0
003	Fat, Crude (%)	0590	1.775	0.1500	1.788	0.2928	0.1156	117	-0.04	003.09	0
003	Fat, Crude (%)	0226	1.780	0.1600	1.788	0.2928	0.1156	117	-0.03	003.09	0
003	Fat, Crude (%)	0187	1.785	0.0300	1.788	0.2928	0.1156	117	-0.01	003.13	0
003	Fat, Crude (%)	0912	1.785	0.0900	1.788	0.2928	0.1156	117	-0.01	003.14	0
003	Fat, Crude (%)	0510	1.800	0.2000	1.788	0.2928	0.1156	117	0.04	003.09	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat, Crude (%)	0148	1.800	0.0200	1.788	0.2928	0.1156	117	0.04	003.10	0
003	Fat, Crude (%)	0889	1.800	0.0600	1.788	0.2928	0.1156	117	0.04	003.10	0
003	Fat, Crude (%)	0644	1.808	0.0140	1.788	0.2928	0.1156	117	0.07	003.06	0
003	Fat, Crude (%)	0536	1.810	0.0600	1.788	0.2928	0.1156	117	0.08	003.99	0
003	Fat, Crude (%)	0098	1.820	0.0600	1.788	0.2928	0.1156	117	0.11	003.09	0
003	Fat, Crude (%)	0720	1.825	0.0300	1.788	0.2928	0.1156	117	0.13	003.10	0
003	Fat, Crude (%)	0123	1.825	0.0300	1.788	0.2928	0.1156	117	0.13	003.13	0
003	Fat, Crude (%)	0881	1.826	0.0760	1.788	0.2928	0.1156	117	0.13	003.10	0
003	Fat, Crude (%)	0674	1.830	0.0000	1.788	0.2928	0.1156	117	0.14	003.09	0
003	Fat, Crude (%)	2196	1.830	0.0000	1.788	0.2928	0.1156	117	0.14	003.10	0
003	Fat, Crude (%)	0884	1.850	0.1000	1.788	0.2928	0.1156	117	0.21	003.06	0
003	Fat, Crude (%)	0723	1.850	0.0000	1.788	0.2928	0.1156	117	0.21	003.09	0
003	Fat, Crude (%)	0689	1.850	0.1000	1.788	0.2928	0.1156	117	0.21	003.10	0
003	Fat, Crude (%)	0692	1.850	0.5000	1.788	0.2928	0.1156	117	0.21	003.14	0
003	Fat, Crude (%)	0164	1.865	0.0700	1.788	0.2928	0.1156	117	0.26	003.00	0
003	Fat, Crude (%)	0870	1.869	0.0881	1.788	0.2928	0.1156	117	0.28	003.09	0
003	Fat, Crude (%)	0037	1.880	0.0000	1.788	0.2928	0.1156	117	0.32	003.14	0
003	Fat, Crude (%)	0066	1.880	0.1200	1.788	0.2928	0.1156	117	0.32	003.14	0
003	Fat, Crude (%)	0885	1.886	0.0231	1.788	0.2928	0.1156	117	0.34	003.00	0
003	Fat, Crude (%)	0968	1.894	0.0190	1.788	0.2928	0.1156	117	0.36	003.14	0
003	Fat, Crude (%)	2006	1.920	0.0600	1.788	0.2928	0.1156	117	0.45	003.13	0
003	Fat, Crude (%)	0687	1.925	0.2500	1.788	0.2928	0.1156	117	0.47	003.06	0
003	Fat, Crude (%)	0868	1.935	0.0100	1.788	0.2928	0.1156	117	0.50	003.14	0
003	Fat, Crude (%)	0049	1.940	0.0600	1.788	0.2928	0.1156	117	0.52	003.14	0
003	Fat, Crude (%)	0676	1.945	0.0100	1.788	0.2928	0.1156	117	0.54	003.10	0
003	Fat, Crude (%)	0229	1.947	0.0030	1.788	0.2928	0.1156	117	0.54	003.06	0
003	Fat, Crude (%)	0689	1.950	0.1000	1.788	0.2928	0.1156	117	0.55	003.06	0
003	Fat, Crude (%)	0619	1.950	0.0400	1.788	0.2928	0.1156	117	0.55	003.14	0
003	Fat, Crude (%)	0353	1.975	0.0100	1.788	0.2928	0.1156	117	0.64	003.09	0
003	Fat, Crude (%)	0630	1.975	0.1700	1.788	0.2928	0.1156	117	0.64	003.09	0
003	Fat, Crude (%)	0653	1.995	0.0300	1.788	0.2928	0.1156	117	0.71	003.00	0
003	Fat, Crude (%)	0918	2.020	0.1200	1.788	0.2928	0.1156	117	0.79	003.06	0
003	Fat, Crude (%)	0682	2.040	0.0000	1.788	0.2928	0.1156	117	0.86	003.06	0
003	Fat, Crude (%)	0229	2.041	0.0440	1.788	0.2928	0.1156	117	0.87	003.14	0
003	Fat, Crude (%)	0278	2.050	0.1000	1.788	0.2928	0.1156	117	0.90	003.14	0
003	Fat, Crude (%)	0683	2.055	0.0300	1.788	0.2928	0.1156	117	0.91	003.14	0
003	Fat, Crude (%)	0646	2.065	0.0100	1.788	0.2928	0.1156	117	0.95	003.13	0
003	Fat, Crude (%)	0309	2.092	0.0156	1.788	0.2928	0.1156	117	1.04	003.00	0
003	Fat, Crude (%)	0910	2.125	0.0300	1.788	0.2928	0.1156	117	1.15	003.06	0
003	Fat, Crude (%)	0148	2.130	0.0200	1.788	0.2928	0.1156	117	1.17	003.14	0
003	Fat, Crude (%)	0870	2.134	0.0904	1.788	0.2928	0.1156	117	1.18	003.13	0
003	Fat, Crude (%)	0413	2.150	0.1000	1.788	0.2928	0.1156	117	1.24	003.14	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat, Crude (%)	0354	2.170	0.0200	1.788	0.2928	0.1156	117	1.31	003.09	0
003	Fat, Crude (%)	2075	2.220	0.3000	1.788	0.2928	0.1156	117	1.48	003.14	0
003	Fat, Crude (%)	2109	2.255	0.4100	1.788	0.2928	0.1156	117	1.60	003.00	0
003	Fat, Crude (%)	0853	2.300	0.6000	1.788	0.2928	0.1156	117	1.75	003.14	0
003	Fat, Crude (%)	0175	2.305	0.0300	1.788	0.2928	0.1156	117	1.77	003.00	0
003	Fat, Crude (%)	0609	2.370	0.0200	1.788	0.2928	0.1156	117	1.99	003.10	0
003	Fat, Crude (%)	0852	2.380	0.0600	1.788	0.2928	0.1156	117	2.02	003.11	0
003	Fat, Crude (%)	0685	2.415	0.0100	1.788	0.2928	0.1156	117	2.14	003.09	0
003	Fat, Crude (%)	0675	2.430	0.1200	1.788	0.2928	0.1156	117	2.19	003.09	0
003	Fat, Crude (%)	0682	2.460	0.0000	1.788	0.2928	0.1156	117	2.30	003.14	0
003	Fat, Crude (%)	2126	2.525	0.0300	1.788	0.2928	0.1156	117	2.52	003.00	0
003	Fat, Crude (%)	0152	2.690	0.0400	1.788	0.2928	0.1156	117	3.08	003.00	0
003	Fat, Crude (%)	0656	2.805	0.4100	1.788	0.2928	0.1156	117	3.47	003.14	0
003	Fat, Crude (%)	0618	3.433	0.0040	1.788	0.2928	0.1156	117	5.62	003.10	0
003	Fat, Crude (%)	2195	6.050	9.500	1.788	0.2928	0.1156	117	14.56	003.10	1
003	Fat, Crude (%)	0638	14.14	0.0400	1.788	0.2928	0.1156	117	42.19	003.09	2
004	Fiber, Crude (%)	2089	8.275	0.2900	11.57	1.592	0.3372	114	-2.07	004.03	0
004	Fiber, Crude (%)	0123	8.670	0.1800	11.57	1.592	0.3372	114	-1.82	004.07	0
004	Fiber, Crude (%)	0028	8.800	0.8000	11.57	1.592	0.3372	114	-1.74	004.07	0
004	Fiber, Crude (%)	0505	8.830	0.1200	11.57	1.592	0.3372	114	-1.72	004.07	0
004	Fiber, Crude (%)	2126	9.015	0.1300	11.57	1.592	0.3372	114	-1.60	004.99	0
004	Fiber, Crude (%)	0619	9.060	0.0000	11.57	1.592	0.3372	114	-1.57	004.07	0
004	Fiber, Crude (%)	0004	9.110	0.4400	11.57	1.592	0.3372	114	-1.54	004.07	0
004	Fiber, Crude (%)	0939	9.260	0.6400	11.57	1.592	0.3372	114	-1.45	004.07	0
004	Fiber, Crude (%)	0683	9.355	0.0700	11.57	1.592	0.3372	114	-1.39	004.07	0
004	Fiber, Crude (%)	0861	9.585	0.0100	11.57	1.592	0.3372	114	-1.25	004.07	0
004	Fiber, Crude (%)	0510	9.650	0.1000	11.57	1.592	0.3372	114	-1.20	004.07	0
004	Fiber, Crude (%)	0610	9.700	0.2000	11.57	1.592	0.3372	114	-1.17	004.99	0
004	Fiber, Crude (%)	0918	9.830	0.0800	11.57	1.592	0.3372	114	-1.09	004.06	0
004	Fiber, Crude (%)	0278	9.850	0.1000	11.57	1.592	0.3372	114	-1.08	004.07	0
004	Fiber, Crude (%)	0956	9.900	1.400	11.57	1.592	0.3372	114	-1.05	004.07	0
004	Fiber, Crude (%)	0682	9.900	0.0000	11.57	1.592	0.3372	114	-1.05	004.07	0
004	Fiber, Crude (%)	0520	10.00	1.400	11.57	1.592	0.3372	114	-0.98	004.07	0
004	Fiber, Crude (%)	0003	10.18	0.5500	11.57	1.592	0.3372	114	-0.87	004.07	0
004	Fiber, Crude (%)	0644	10.20	0.5120	11.57	1.592	0.3372	114	-0.86	004.07	0
004	Fiber, Crude (%)	0019	10.22	0.0800	11.57	1.592	0.3372	114	-0.85	004.07	0
004	Fiber, Crude (%)	0848	10.28	0.0000	11.57	1.592	0.3372	114	-0.81	004.07	0
004	Fiber, Crude (%)	2190	10.32	0.9500	11.57	1.592	0.3372	114	-0.79	004.07	0
004	Fiber, Crude (%)	0849	10.32	0.1800	11.57	1.592	0.3372	114	-0.78	004.07	0
004	Fiber, Crude (%)	0100	10.35	0.1300	11.57	1.592	0.3372	114	-0.77	004.07	0
004	Fiber, Crude (%)	0265	10.35	0.5000	11.57	1.592	0.3372	114	-0.76	004.07	0
004	Fiber, Crude (%)	0675	10.43	0.0300	11.57	1.592	0.3372	114	-0.72	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber, Crude (%)	0885	10.47	0.1671	11.57	1.592	0.3372	114	-0.69	004.06	0
004	Fiber, Crude (%)	2109	10.49	0.4400	11.57	1.592	0.3372	114	-0.68	004.07	0
004	Fiber, Crude (%)	0190	10.50	0.2000	11.57	1.592	0.3372	114	-0.67	004.07	0
004	Fiber, Crude (%)	0948	10.53	0.2700	11.57	1.592	0.3372	114	-0.65	004.07	0
004	Fiber, Crude (%)	0695	10.59	0.6500	11.57	1.592	0.3372	114	-0.62	004.00	0
004	Fiber, Crude (%)	2193	10.60	0.1700	11.57	1.592	0.3372	114	-0.61	004.07	0
004	Fiber, Crude (%)	0345	10.62	0.2300	11.57	1.592	0.3372	114	-0.60	004.00	0
004	Fiber, Crude (%)	0074	10.64	0.4400	11.57	1.592	0.3372	114	-0.58	004.07	0
004	Fiber, Crude (%)	0354	10.72	0.0400	11.57	1.592	0.3372	114	-0.53	004.00	0
004	Fiber, Crude (%)	2145	10.78	0.2000	11.57	1.592	0.3372	114	-0.49	004.99	0
004	Fiber, Crude (%)	0968	10.81	0.3820	11.57	1.592	0.3372	114	-0.48	004.06	0
004	Fiber, Crude (%)	0692	10.85	0.7000	11.57	1.592	0.3372	114	-0.45	004.07	0
004	Fiber, Crude (%)	0968	10.92	0.1340	11.57	1.592	0.3372	114	-0.41	004.07	0
004	Fiber, Crude (%)	0675	10.95	0.2700	11.57	1.592	0.3372	114	-0.39	004.06	0
004	Fiber, Crude (%)	2006	10.97	0.1400	11.57	1.592	0.3372	114	-0.38	004.06	0
004	Fiber, Crude (%)	0169	10.99	0.0900	11.57	1.592	0.3372	114	-0.37	004.00	0
004	Fiber, Crude (%)	0309	11.10	0.3071	11.57	1.592	0.3372	114	-0.29	004.00	0
004	Fiber, Crude (%)	0720	11.25	0.1000	11.57	1.592	0.3372	114	-0.20	004.11	0
004	Fiber, Crude (%)	0904	11.27	0.5900	11.57	1.592	0.3372	114	-0.19	004.03	0
004	Fiber, Crude (%)	0098	11.31	0.7500	11.57	1.592	0.3372	114	-0.16	004.07	0
004	Fiber, Crude (%)	0596	11.32	0.0400	11.57	1.592	0.3372	114	-0.16	004.07	0
004	Fiber, Crude (%)	0656	11.33	0.5200	11.57	1.592	0.3372	114	-0.15	004.07	0
004	Fiber, Crude (%)	2009	11.35	0.0553	11.57	1.592	0.3372	114	-0.14	004.07	0
004	Fiber, Crude (%)	2073	11.36	0.5300	11.57	1.592	0.3372	114	-0.13	004.00	0
004	Fiber, Crude (%)	0353	11.36	0.1100	11.57	1.592	0.3372	114	-0.13	004.03	0
004	Fiber, Crude (%)	2190	11.37	0.5300	11.57	1.592	0.3372	114	-0.13	004.03	0
004	Fiber, Crude (%)	0027	11.40	0.1150	11.57	1.592	0.3372	114	-0.10	004.06	0
004	Fiber, Crude (%)	0164	11.45	0.1000	11.57	1.592	0.3372	114	-0.07	004.00	0
004	Fiber, Crude (%)	0171	11.45	0.3000	11.57	1.592	0.3372	114	-0.07	004.00	0
004	Fiber, Crude (%)	0897	11.50	0.0200	11.57	1.592	0.3372	114	-0.04	004.06	0
004	Fiber, Crude (%)	0964	11.51	0.5400	11.57	1.592	0.3372	114	-0.04	004.03	0
004	Fiber, Crude (%)	2192	11.52	0.0100	11.57	1.592	0.3372	114	-0.03	004.03	0
004	Fiber, Crude (%)	0653	11.52	0.4400	11.57	1.592	0.3372	114	-0.03	004.07	0
004	Fiber, Crude (%)	2022	11.56	0.0500	11.57	1.592	0.3372	114	-0.01	004.06	0
004	Fiber, Crude (%)	0876	11.57	0.0535	11.57	1.592	0.3372	114	0.00	004.07	0
004	Fiber, Crude (%)	2196	11.57	0.0000	11.57	1.592	0.3372	114	0.00	004.07	0
004	Fiber, Crude (%)	0205	11.60	0.4000	11.57	1.592	0.3372	114	0.02	004.06	0
004	Fiber, Crude (%)	0726	11.65	0.0500	11.57	1.592	0.3372	114	0.05	004.07	0
004	Fiber, Crude (%)	0263	11.66	0.2177	11.57	1.592	0.3372	114	0.06	004.07	0
004	Fiber, Crude (%)	0689	11.70	0.4000	11.57	1.592	0.3372	114	0.08	004.06	0
004	Fiber, Crude (%)	0881	11.74	0.5670	11.57	1.592	0.3372	114	0.11	004.06	0
004	Fiber, Crude (%)	0202	11.75	0.8530	11.57	1.592	0.3372	114	0.12	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber, Crude (%)	0425	11.85	0.1000	11.57	1.592	0.3372	114	0.18	004.00	0
004	Fiber, Crude (%)	2076	11.88	0.1728	11.57	1.592	0.3372	114	0.19	004.00	0
004	Fiber, Crude (%)	2198	11.89	0.1800	11.57	1.592	0.3372	114	0.20	004.06	0
004	Fiber, Crude (%)	0868	11.90	0.0600	11.57	1.592	0.3372	114	0.21	004.06	0
004	Fiber, Crude (%)	0098	11.94	0.4500	11.57	1.592	0.3372	114	0.23	004.06	0
004	Fiber, Crude (%)	0952	11.94	0.0800	11.57	1.592	0.3372	114	0.23	004.06	0
004	Fiber, Crude (%)	0175	11.95	0.1000	11.57	1.592	0.3372	114	0.24	004.00	0
004	Fiber, Crude (%)	0674	11.98	0.0000	11.57	1.592	0.3372	114	0.26	004.06	0
004	Fiber, Crude (%)	0229	12.00	0.2200	11.57	1.592	0.3372	114	0.27	004.07	0
004	Fiber, Crude (%)	0893	12.02	0.0000	11.57	1.592	0.3372	114	0.28	004.06	0
004	Fiber, Crude (%)	0941	12.03	0.1200	11.57	1.592	0.3372	114	0.29	004.06	0
004	Fiber, Crude (%)	0089	12.10	0.0200	11.57	1.592	0.3372	114	0.33	004.07	0
004	Fiber, Crude (%)	0609	12.15	0.1700	11.57	1.592	0.3372	114	0.36	004.06	0
004	Fiber, Crude (%)	0638	12.15	0.1000	11.57	1.592	0.3372	114	0.37	004.06	0
004	Fiber, Crude (%)	0646	12.16	1.470	11.57	1.592	0.3372	114	0.37	004.07	0
004	Fiber, Crude (%)	0870	12.22	0.3251	11.57	1.592	0.3372	114	0.41	004.07	0
004	Fiber, Crude (%)	0723	12.27	0.0000	11.57	1.592	0.3372	114	0.44	004.07	0
004	Fiber, Crude (%)	0511	12.30	0.2000	11.57	1.592	0.3372	114	0.46	004.00	0
004	Fiber, Crude (%)	0878	12.47	0.0500	11.57	1.592	0.3372	114	0.56	004.06	0
004	Fiber, Crude (%)	0009	12.51	0.2600	11.57	1.592	0.3372	114	0.59	004.07	0
004	Fiber, Crude (%)	2062	12.52	0.2107	11.57	1.592	0.3372	114	0.60	004.06	0
004	Fiber, Crude (%)	0948	12.58	0.3100	11.57	1.592	0.3372	114	0.63	004.06	0
004	Fiber, Crude (%)	0884	12.65	0.7000	11.57	1.592	0.3372	114	0.68	004.00	0
004	Fiber, Crude (%)	2181	12.76	0.0800	11.57	1.592	0.3372	114	0.75	004.07	0
004	Fiber, Crude (%)	0910	12.80	0.8000	11.57	1.592	0.3372	114	0.77	004.07	0
004	Fiber, Crude (%)	0407	13.17	0.7694	11.57	1.592	0.3372	114	1.01	004.07	0
004	Fiber, Crude (%)	0720	13.38	0.3700	11.57	1.592	0.3372	114	1.14	004.06	0
004	Fiber, Crude (%)	2188	13.59	1.500	11.57	1.592	0.3372	114	1.27	004.07	0
004	Fiber, Crude (%)	0083	13.61	0.8400	11.57	1.592	0.3372	114	1.28	004.07	0
004	Fiber, Crude (%)	0889	13.66	0.7300	11.57	1.592	0.3372	114	1.31	004.11	0
004	Fiber, Crude (%)	0144	13.80	0.1800	11.57	1.592	0.3372	114	1.40	004.07	0
004	Fiber, Crude (%)	0405	13.80	0.0000	11.57	1.592	0.3372	114	1.40	004.11	0
004	Fiber, Crude (%)	0504	13.89	0.1900	11.57	1.592	0.3372	114	1.46	004.00	0
004	Fiber, Crude (%)	0852	13.89	0.2900	11.57	1.592	0.3372	114	1.46	004.11	0
004	Fiber, Crude (%)	0626	14.03	0.9200	11.57	1.592	0.3372	114	1.55	004.07	0
004	Fiber, Crude (%)	0536	14.16	1.440	11.57	1.592	0.3372	114	1.63	004.07	0
004	Fiber, Crude (%)	0016	14.40	0.4000	11.57	1.592	0.3372	114	1.78	004.07	0
004	Fiber, Crude (%)	0598	14.45	0.4300	11.57	1.592	0.3372	114	1.81	004.07	0
004	Fiber, Crude (%)	0226	14.61	0.0400	11.57	1.592	0.3372	114	1.91	004.00	0
004	Fiber, Crude (%)	0354	14.69	0.1300	11.57	1.592	0.3372	114	1.96	004.07	0
004	Fiber, Crude (%)	0693	14.72	0.3600	11.57	1.592	0.3372	114	1.98	004.03	0
004	Fiber, Crude (%)	0034	15.07	0.0700	11.57	1.592	0.3372	114	2.20	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
004	Fiber, Crude (%)	0032	15.23	0.8400	11.57	1.592	0.3372	114	2.30	004.07	0
004	Fiber, Crude (%)	0955	15.25	0.9000	11.57	1.592	0.3372	114	2.31	004.07	0
004	Fiber, Crude (%)	0066	15.40	0.5000	11.57	1.592	0.3372	114	2.41	004.07	0
004	Fiber, Crude (%)	0045	17.35	0.3000	11.57	1.592	0.3372	114	3.63	004.03	0
004	Fiber, Crude (%)	2200	8.950	2.100	11.57	1.592	0.3372	114	-1.64	004.06	1
004	Fiber, Crude (%)	0553	14.78	2.240	11.57	1.592	0.3372	114	2.02	004.07	1
004	Fiber, Crude (%)	0643	0.1297	0.0052	11.57	1.592	0.3372	114	-7.18	004.07	2
005	Ash (%)	0720	13.59	0.1000	20.84	0.2733	0.1330	137	-26.53	005.11	0
005	Ash (%)	0853	20.00	1.800	20.84	0.2733	0.1330	137	-3.08	005.00	0
005	Ash (%)	0674	20.26	0.0000	20.84	0.2733	0.1330	137	-2.13	005.00	0
005	Ash (%)	2081	20.30	0.3500	20.84	0.2733	0.1330	137	-2.00	005.99	0
005	Ash (%)	0539	20.37	0.0800	20.84	0.2733	0.1330	137	-1.73	005.00	0
005	Ash (%)	0425	20.39	0.0200	20.84	0.2733	0.1330	137	-1.65	005.00	0
005	Ash (%)	0169	20.43	0.0200	20.84	0.2733	0.1330	137	-1.51	005.00	0
005	Ash (%)	0511	20.45	0.1000	20.84	0.2733	0.1330	137	-1.43	005.00	0
005	Ash (%)	0874	20.50	0.0300	20.84	0.2733	0.1330	137	-1.27	005.00	0
005	Ash (%)	0108	20.50	0.1400	20.84	0.2733	0.1330	137	-1.25	005.00	0
005	Ash (%)	0175	20.50	0.0000	20.84	0.2733	0.1330	137	-1.25	005.00	0
005	Ash (%)	0164	20.52	0.0100	20.84	0.2733	0.1330	137	-1.19	005.00	0
005	Ash (%)	0726	20.53	0.0300	20.84	0.2733	0.1330	137	-1.16	005.05	0
005	Ash (%)	0589	20.54	0.0300	20.84	0.2733	0.1330	137	-1.12	005.00	0
005	Ash (%)	2006	20.55	0.0000	20.84	0.2733	0.1330	137	-1.07	005.00	0
005	Ash (%)	0870	20.56	0.0567	20.84	0.2733	0.1330	137	-1.04	005.00	0
005	Ash (%)	0618	20.56	0.0566	20.84	0.2733	0.1330	137	-1.04	005.00	0
005	Ash (%)	0004	20.58	0.0100	20.84	0.2733	0.1330	137	-0.97	005.00	0
005	Ash (%)	0596	20.58	0.0700	20.84	0.2733	0.1330	137	-0.97	005.00	0
005	Ash (%)	2109	20.58	0.0100	20.84	0.2733	0.1330	137	-0.97	005.05	0
005	Ash (%)	0278	20.58	0.3000	20.84	0.2733	0.1330	137	-0.96	005.00	0
005	Ash (%)	0682	20.59	0.0000	20.84	0.2733	0.1330	137	-0.92	005.00	0
005	Ash (%)	0358	20.60	0.0400	20.84	0.2733	0.1330	137	-0.88	005.00	0
005	Ash (%)	2223	20.61	0.0500	20.84	0.2733	0.1330	137	-0.87	005.00	0
005	Ash (%)	0019	20.61	0.1600	20.84	0.2733	0.1330	137	-0.85	005.00	0
005	Ash (%)	0171	20.62	0.0300	20.84	0.2733	0.1330	137	-0.83	005.00	0
005	Ash (%)	0651	20.62	0.0710	20.84	0.2733	0.1330	137	-0.82	005.00	0
005	Ash (%)	0066	20.62	0.0600	20.84	0.2733	0.1330	137	-0.81	005.05	0
005	Ash (%)	2150	20.63	0.0100	20.84	0.2733	0.1330	137	-0.79	005.00	0
005	Ash (%)	2193	20.63	0.0700	20.84	0.2733	0.1330	137	-0.79	005.00	0
005	Ash (%)	0541	20.65	0.0300	20.84	0.2733	0.1330	137	-0.72	005.00	0
005	Ash (%)	0878	20.65	0.1100	20.84	0.2733	0.1330	137	-0.72	005.05	0
005	Ash (%)	0265	20.65	0.1000	20.84	0.2733	0.1330	137	-0.70	005.00	0
005	Ash (%)	0357	20.65	0.1000	20.84	0.2733	0.1330	137	-0.70	005.00	0
005	Ash (%)	0049	20.66	0.0500	20.84	0.2733	0.1330	137	-0.68	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0144	20.66	0.0300	20.84	0.2733	0.1330	137	-0.68	005.00	0
005	Ash (%)	0646	20.67	0.0700	20.84	0.2733	0.1330	137	-0.65	005.00	0
005	Ash (%)	2181	20.67	0.1700	20.84	0.2733	0.1330	137	-0.65	005.00	0
005	Ash (%)	0695	20.67	0.0610	20.84	0.2733	0.1330	137	-0.62	005.05	0
005	Ash (%)	0263	20.68	0.0016	20.84	0.2733	0.1330	137	-0.59	005.05	0
005	Ash (%)	2192	20.69	0.0200	20.84	0.2733	0.1330	137	-0.55	005.05	0
005	Ash (%)	0723	20.69	0.0000	20.84	0.2733	0.1330	137	-0.55	005.00	0
005	Ash (%)	0510	20.70	0.0100	20.84	0.2733	0.1330	137	-0.54	005.00	0
005	Ash (%)	0574	20.70	0.0100	20.84	0.2733	0.1330	137	-0.54	005.99	0
005	Ash (%)	0027	20.70	0.0380	20.84	0.2733	0.1330	137	-0.53	005.00	0
005	Ash (%)	0001	20.70	0.0068	20.84	0.2733	0.1330	137	-0.52	005.00	0
005	Ash (%)	0653	20.70	0.0000	20.84	0.2733	0.1330	137	-0.52	005.00	0
005	Ash (%)	0912	20.70	0.0000	20.84	0.2733	0.1330	137	-0.52	005.03	0
005	Ash (%)	0186	20.70	0.2000	20.84	0.2733	0.1330	137	-0.52	005.00	0
005	Ash (%)	0520	20.72	0.1300	20.84	0.2733	0.1330	137	-0.46	005.00	0
005	Ash (%)	0205	20.73	0.0500	20.84	0.2733	0.1330	137	-0.43	005.00	0
005	Ash (%)	2195	20.74	0.0000	20.84	0.2733	0.1330	137	-0.37	005.00	0
005	Ash (%)	0861	20.74	0.0200	20.84	0.2733	0.1330	137	-0.37	005.00	0
005	Ash (%)	0354	20.75	0.0000	20.84	0.2733	0.1330	137	-0.33	005.00	0
005	Ash (%)	2200	20.75	0.1000	20.84	0.2733	0.1330	137	-0.33	005.00	0
005	Ash (%)	0692	20.75	0.3000	20.84	0.2733	0.1330	137	-0.33	005.05	0
005	Ash (%)	0152	20.77	0.0700	20.84	0.2733	0.1330	137	-0.28	005.05	0
005	Ash (%)	0505	20.77	0.0600	20.84	0.2733	0.1330	137	-0.26	005.00	0
005	Ash (%)	0536	20.78	0.0700	20.84	0.2733	0.1330	137	-0.24	005.99	0
005	Ash (%)	0407	20.78	0.0394	20.84	0.2733	0.1330	137	-0.24	005.00	0
005	Ash (%)	0098	20.78	0.3000	20.84	0.2733	0.1330	137	-0.22	005.00	0
005	Ash (%)	0656	20.78	0.1600	20.84	0.2733	0.1330	137	-0.22	005.00	0
005	Ash (%)	0689	20.79	0.0100	20.84	0.2733	0.1330	137	-0.21	005.05	0
005	Ash (%)	0849	20.80	0.0090	20.84	0.2733	0.1330	137	-0.17	005.05	0
005	Ash (%)	0045	20.80	0.2000	20.84	0.2733	0.1330	137	-0.15	005.00	0
005	Ash (%)	0142	20.80	0.0400	20.84	0.2733	0.1330	137	-0.15	005.00	0
005	Ash (%)	0202	20.80	0.0340	20.84	0.2733	0.1330	137	-0.15	005.99	0
005	Ash (%)	0553	20.81	0.0100	20.84	0.2733	0.1330	137	-0.13	005.00	0
005	Ash (%)	0953	20.81	0.1100	20.84	0.2733	0.1330	137	-0.13	005.00	0
005	Ash (%)	0504	20.82	0.1500	20.84	0.2733	0.1330	137	-0.10	005.00	0
005	Ash (%)	2188	20.82	0.0100	20.84	0.2733	0.1330	137	-0.10	005.00	0
005	Ash (%)	0187	20.83	0.0200	20.84	0.2733	0.1330	137	-0.04	005.00	0
005	Ash (%)	0848	20.84	0.0100	20.84	0.2733	0.1330	137	-0.02	005.00	0
005	Ash (%)	0083	20.85	0.4800	20.84	0.2733	0.1330	137	0.03	005.00	0
005	Ash (%)	0956	20.85	0.1000	20.84	0.2733	0.1330	137	0.03	005.00	0
005	Ash (%)	2022	20.85	0.0000	20.84	0.2733	0.1330	137	0.03	005.00	0
005	Ash (%)	0003	20.85	0.1000	20.84	0.2733	0.1330	137	0.03	005.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0893	20.85	0.0000	20.84	0.2733	0.1330	137	0.03	005.05	0
005	Ash (%)	0598	20.86	0.0300	20.84	0.2733	0.1330	137	0.05	005.00	0
005	Ash (%)	2076	20.86	0.0945	20.84	0.2733	0.1330	137	0.08	005.00	0
005	Ash (%)	0016	20.90	0.0000	20.84	0.2733	0.1330	137	0.21	005.00	0
005	Ash (%)	0619	20.90	0.0000	20.84	0.2733	0.1330	137	0.21	005.05	0
005	Ash (%)	0876	20.90	0.5575	20.84	0.2733	0.1330	137	0.21	005.00	0
005	Ash (%)	0226	20.91	0.2100	20.84	0.2733	0.1330	137	0.23	005.00	0
005	Ash (%)	2053	20.91	0.0800	20.84	0.2733	0.1330	137	0.25	005.00	0
005	Ash (%)	0948	20.92	0.0300	20.84	0.2733	0.1330	137	0.27	005.00	0
005	Ash (%)	0683	20.92	0.4700	20.84	0.2733	0.1330	137	0.27	005.05	0
005	Ash (%)	2012	20.92	0.1900	20.84	0.2733	0.1330	137	0.27	005.05	0
005	Ash (%)	0345	20.94	0.0700	20.84	0.2733	0.1330	137	0.34	005.00	0
005	Ash (%)	2215	20.94	0.0000	20.84	0.2733	0.1330	137	0.36	005.99	0
005	Ash (%)	0590	20.95	0.1000	20.84	0.2733	0.1330	137	0.40	005.05	0
005	Ash (%)	0309	20.96	0.0590	20.84	0.2733	0.1330	137	0.42	005.00	0
005	Ash (%)	0630	20.97	0.2900	20.84	0.2733	0.1330	137	0.45	005.00	0
005	Ash (%)	0881	20.97	0.1100	20.84	0.2733	0.1330	137	0.45	005.00	0
005	Ash (%)	0638	20.97	0.0600	20.84	0.2733	0.1330	137	0.47	005.05	0
005	Ash (%)	0353	20.98	0.0300	20.84	0.2733	0.1330	137	0.49	005.00	0
005	Ash (%)	2196	20.98	0.0000	20.84	0.2733	0.1330	137	0.51	005.00	0
005	Ash (%)	0629	20.99	0.2100	20.84	0.2733	0.1330	137	0.53	005.99	0
005	Ash (%)	0889	20.99	0.2400	20.84	0.2733	0.1330	137	0.54	005.05	0
005	Ash (%)	0968	21.00	0.0200	20.84	0.2733	0.1330	137	0.57	005.00	0
005	Ash (%)	2062	21.00	0.2000	20.84	0.2733	0.1330	137	0.58	005.05	0
005	Ash (%)	0918	21.01	0.0800	20.84	0.2733	0.1330	137	0.62	005.00	0
005	Ash (%)	0897	21.01	0.0000	20.84	0.2733	0.1330	137	0.62	005.05	0
005	Ash (%)	0868	21.03	0.0100	20.84	0.2733	0.1330	137	0.67	005.05	0
005	Ash (%)	0229	21.03	0.1400	20.84	0.2733	0.1330	137	0.69	005.00	0
005	Ash (%)	0885	21.03	0.0373	20.84	0.2733	0.1330	137	0.69	005.05	0
005	Ash (%)	0884	21.05	0.1000	20.84	0.2733	0.1330	137	0.76	005.99	0
005	Ash (%)	0100	21.06	0.5600	20.84	0.2733	0.1330	137	0.80	005.00	0
005	Ash (%)	2009	21.07	0.0086	20.84	0.2733	0.1330	137	0.83	005.05	0
005	Ash (%)	0720	21.07	0.0400	20.84	0.2733	0.1330	137	0.84	005.05	0
005	Ash (%)	0119	21.08	0.1200	20.84	0.2733	0.1330	137	0.87	005.00	0
005	Ash (%)	0609	21.10	0.1500	20.84	0.2733	0.1330	137	0.93	005.00	0
005	Ash (%)	0610	21.11	0.0500	20.84	0.2733	0.1330	137	0.96	005.05	0
005	Ash (%)	0852	21.13	0.1100	20.84	0.2733	0.1330	137	1.04	005.00	0
005	Ash (%)	2126	21.13	0.0100	20.84	0.2733	0.1330	137	1.04	005.05	0
005	Ash (%)	0693	21.14	0.1520	20.84	0.2733	0.1330	137	1.09	005.00	0
005	Ash (%)	0047	21.15	0.1000	20.84	0.2733	0.1330	137	1.13	005.00	0
005	Ash (%)	2190	21.15	0.2400	20.84	0.2733	0.1330	137	1.13	005.05	0
005	Ash (%)	0675	21.16	0.0900	20.84	0.2733	0.1330	137	1.15	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	2198	21.17	0.0600	20.84	0.2733	0.1330	137	1.20	005.05	0
005	Ash (%)	2073	21.21	0.3700	20.84	0.2733	0.1330	137	1.33	005.00	0
005	Ash (%)	0123	21.25	0.5000	20.84	0.2733	0.1330	137	1.49	005.00	0
005	Ash (%)	0685	21.29	0.0000	20.84	0.2733	0.1330	137	1.64	005.05	0
005	Ash (%)	0148	21.31	0.1900	20.84	0.2733	0.1330	137	1.70	005.00	0
005	Ash (%)	2145	21.31	0.1800	20.84	0.2733	0.1330	137	1.71	005.05	0
005	Ash (%)	0880	21.33	0.6500	20.84	0.2733	0.1330	137	1.77	005.00	0
005	Ash (%)	0676	21.38	0.0300	20.84	0.2733	0.1330	137	1.95	005.99	0
005	Ash (%)	2146	21.38	0.2200	20.84	0.2733	0.1330	137	1.97	005.05	0
005	Ash (%)	0939	21.39	0.0800	20.84	0.2733	0.1330	137	2.01	005.99	0
005	Ash (%)	0644	21.40	0.3000	20.84	0.2733	0.1330	137	2.04	005.02	0
005	Ash (%)	0034	21.43	0.6580	20.84	0.2733	0.1330	137	2.15	005.00	0
005	Ash (%)	0955	21.45	0.9000	20.84	0.2733	0.1330	137	2.23	005.05	0
005	Ash (%)	0941	21.48	0.1600	20.84	0.2733	0.1330	137	2.34	005.05	0
005	Ash (%)	0190	21.49	0.0600	20.84	0.2733	0.1330	137	2.37	005.05	0
005	Ash (%)	0910	21.55	0.3000	20.84	0.2733	0.1330	137	2.59	005.99	0
005	Ash (%)	2129	24.35	0.5000	20.84	0.2733	0.1330	137	12.84	005.05	0
005	Ash (%)	0889	29.37	0.7200	20.84	0.2733	0.1330	137	31.21	005.11	0
005	Ash (%)	2089	25.42	10.15	20.84	0.2733	0.1330	137	16.74	005.00	1
005	Ash (%)	0643	0.2079	0.0018	20.84	0.2733	0.1330	137	-75.50	005.00	2
005	Ash (%)	0852	2.085	0.0100	20.84	0.2733	0.1330	137	-68.63	005.11	2
006	Total Sugars (%)	0148	3.450	0.3000	4.793	1.426	0.2746	5		006.99	0
006	Total Sugars (%)	0910	3.885	0.0100	4.793	1.426	0.2746	5		006.00	0
006	Total Sugars (%)	0956	4.600	0.2000	4.793	1.426	0.2746	5		006.99	0
006	Total Sugars (%)	0407	4.899	0.5430	4.793	1.426	0.2746	5		006.00	0
006	Total Sugars (%)	2190	7.130	0.3200	4.793	1.426	0.2746	5		006.01	0
006	Total Sugars (%)	0720	0.0000	0.0000	4.793	1.426	0.2746	5		006.99	4
008	Fiber, Acid Detergent (%)	0004	11.03	0.2700	14.32	2.226	0.3892	59	-1.48	008.08	0
008	Fiber, Acid Detergent (%)	0164	11.25	0.3000	14.32	2.226	0.3892	59	-1.38	008.08	0
008	Fiber, Acid Detergent (%)	0968	11.34	0.1530	14.32	2.226	0.3892	59	-1.34	008.08	0
008	Fiber, Acid Detergent (%)	0956	11.45	0.1000	14.32	2.226	0.3892	59	-1.29	008.08	0
008	Fiber, Acid Detergent (%)	0948	11.49	0.1600	14.32	2.226	0.3892	59	-1.27	008.02	0
008	Fiber, Acid Detergent (%)	0939	11.71	1.340	14.32	2.226	0.3892	59	-1.17	008.08	0
008	Fiber, Acid Detergent (%)	0861	11.76	0.1900	14.32	2.226	0.3892	59	-1.15	008.08	0
008	Fiber, Acid Detergent (%)	0910	12.26	0.5000	14.32	2.226	0.3892	59	-0.92	008.08	0
008	Fiber, Acid Detergent (%)	0878	12.27	0.1500	14.32	2.226	0.3892	59	-0.92	008.08	0
008	Fiber, Acid Detergent (%)	0049	12.33	0.3300	14.32	2.226	0.3892	59	-0.90	008.08	0
008	Fiber, Acid Detergent (%)	0278	12.40	0.8000	14.32	2.226	0.3892	59	-0.86	008.08	0
008	Fiber, Acid Detergent (%)	2190	12.60	0.5700	14.32	2.226	0.3892	59	-0.77	008.08	0
008	Fiber, Acid Detergent (%)	0646	12.66	1.230	14.32	2.226	0.3892	59	-0.75	008.08	0
008	Fiber, Acid Detergent (%)	0590	12.75	0.3000	14.32	2.226	0.3892	59	-0.70	008.02	0
008	Fiber, Acid Detergent (%)	0148	12.83	0.9200	14.32	2.226	0.3892	59	-0.67	008.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
008	Fiber, Acid Detergent (%)	0918	12.86	0.4316	14.32	2.226	0.3892	59	-0.65	008.08	0
008	Fiber, Acid Detergent (%)	0510	12.95	0.5000	14.32	2.226	0.3892	59	-0.61	008.08	0
008	Fiber, Acid Detergent (%)	0357	13.00	0.8000	14.32	2.226	0.3892	59	-0.59	008.08	0
008	Fiber, Acid Detergent (%)	0358	13.11	0.3000	14.32	2.226	0.3892	59	-0.54	008.08	0
008	Fiber, Acid Detergent (%)	0066	13.17	0.6500	14.32	2.226	0.3892	59	-0.52	008.08	0
008	Fiber, Acid Detergent (%)	0263	13.29	0.0040	14.32	2.226	0.3892	59	-0.46	008.08	0
008	Fiber, Acid Detergent (%)	2129	13.30	0.7500	14.32	2.226	0.3892	59	-0.46	008.08	0
008	Fiber, Acid Detergent (%)	0848	13.34	0.0000	14.32	2.226	0.3892	59	-0.44	008.08	0
008	Fiber, Acid Detergent (%)	0653	13.35	0.2700	14.32	2.226	0.3892	59	-0.44	008.08	0
008	Fiber, Acid Detergent (%)	2196	13.45	0.0000	14.32	2.226	0.3892	59	-0.39	008.08	0
008	Fiber, Acid Detergent (%)	0354	13.46	0.0500	14.32	2.226	0.3892	59	-0.39	008.08	0
008	Fiber, Acid Detergent (%)	0504	13.54	0.8400	14.32	2.226	0.3892	59	-0.35	008.02	0
008	Fiber, Acid Detergent (%)	0870	13.65	0.7243	14.32	2.226	0.3892	59	-0.30	008.08	0
008	Fiber, Acid Detergent (%)	2193	13.67	0.4800	14.32	2.226	0.3892	59	-0.29	008.08	0
008	Fiber, Acid Detergent (%)	0904	13.76	0.5700	14.32	2.226	0.3892	59	-0.25	008.08	0
008	Fiber, Acid Detergent (%)	0407	14.08	0.0452	14.32	2.226	0.3892	59	-0.11	008.08	0
008	Fiber, Acid Detergent (%)	0619	14.20	0.6000	14.32	2.226	0.3892	59	-0.05	008.08	0
008	Fiber, Acid Detergent (%)	0889	14.29	0.0700	14.32	2.226	0.3892	59	-0.01	008.08	0
008	Fiber, Acid Detergent (%)	2192	14.47	0.0200	14.32	2.226	0.3892	59	0.07	008.02	0
008	Fiber, Acid Detergent (%)	0941	14.83	0.1600	14.32	2.226	0.3892	59	0.23	008.99	0
008	Fiber, Acid Detergent (%)	0609	14.84	0.0100	14.32	2.226	0.3892	59	0.23	008.02	0
008	Fiber, Acid Detergent (%)	0885	14.87	0.4704	14.32	2.226	0.3892	59	0.25	008.02	0
008	Fiber, Acid Detergent (%)	0098	15.26	0.0600	14.32	2.226	0.3892	59	0.42	008.02	0
008	Fiber, Acid Detergent (%)	0674	15.33	0.0000	14.32	2.226	0.3892	59	0.45	008.02	0
008	Fiber, Acid Detergent (%)	0689	15.40	0.6000	14.32	2.226	0.3892	59	0.49	008.02	0
008	Fiber, Acid Detergent (%)	0511	15.40	0.2000	14.32	2.226	0.3892	59	0.49	008.99	0
008	Fiber, Acid Detergent (%)	0345	15.53	0.2500	14.32	2.226	0.3892	59	0.54	008.02	0
008	Fiber, Acid Detergent (%)	0353	15.86	0.1600	14.32	2.226	0.3892	59	0.69	008.02	0
008	Fiber, Acid Detergent (%)	0226	16.04	0.0400	14.32	2.226	0.3892	59	0.77	008.02	0
008	Fiber, Acid Detergent (%)	0016	16.05	0.1000	14.32	2.226	0.3892	59	0.78	008.08	0
008	Fiber, Acid Detergent (%)	0148	16.11	0.2600	14.32	2.226	0.3892	59	0.80	008.02	0
008	Fiber, Acid Detergent (%)	0675	16.23	0.1800	14.32	2.226	0.3892	59	0.86	008.02	0
008	Fiber, Acid Detergent (%)	0035	16.24	0.1700	14.32	2.226	0.3892	59	0.86	008.99	0
008	Fiber, Acid Detergent (%)	0187	16.36	0.3800	14.32	2.226	0.3892	59	0.92	008.02	0
008	Fiber, Acid Detergent (%)	0265	16.45	0.7000	14.32	2.226	0.3892	59	0.96	008.05	0
008	Fiber, Acid Detergent (%)	0001	16.91	0.5000	14.32	2.226	0.3892	59	1.16	008.08	0
008	Fiber, Acid Detergent (%)	0309	17.08	1.801	14.32	2.226	0.3892	59	1.24	008.02	0
008	Fiber, Acid Detergent (%)	0726	17.31	0.5300	14.32	2.226	0.3892	59	1.34	008.08	0
008	Fiber, Acid Detergent (%)	0083	17.47	0.4600	14.32	2.226	0.3892	59	1.42	008.08	0
008	Fiber, Acid Detergent (%)	0190	17.56	0.1500	14.32	2.226	0.3892	59	1.45	008.08	0
008	Fiber, Acid Detergent (%)	0037	17.74	0.1000	14.32	2.226	0.3892	59	1.54	008.08	0
008	Fiber, Acid Detergent (%)	0720	18.26	0.0100	14.32	2.226	0.3892	59	1.77	008.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
008	Fiber, Acid Detergent (%)	0045	18.35	1.100	14.32	2.226	0.3892	59	1.81	008.08	0
008	Fiber, Acid Detergent (%)	0693	18.56	0.1560	14.32	2.226	0.3892	59	1.90	008.08	0
008	Fiber, Acid Detergent (%)	0536	13.71	2.530	14.32	2.226	0.3892	59	-0.28	008.08	1
009	Fiber, Neutral Detergent (%)	0164	20.05	0.1000	24.85	2.538	0.4090	58	-1.89	009.09	0
009	Fiber, Neutral Detergent (%)	0870	20.20	0.2768	24.85	2.538	0.4090	58	-1.83	009.09	0
009	Fiber, Neutral Detergent (%)	0049	21.00	0.7300	24.85	2.538	0.4090	58	-1.52	009.09	0
009	Fiber, Neutral Detergent (%)	0939	21.15	0.5700	24.85	2.538	0.4090	58	-1.46	009.09	0
009	Fiber, Neutral Detergent (%)	0861	21.29	0.6200	24.85	2.538	0.4090	58	-1.40	009.09	0
009	Fiber, Neutral Detergent (%)	0968	21.72	0.2910	24.85	2.538	0.4090	58	-1.23	009.09	0
009	Fiber, Neutral Detergent (%)	0956	21.80	0.4000	24.85	2.538	0.4090	58	-1.20	009.09	0
009	Fiber, Neutral Detergent (%)	0878	21.93	0.7200	24.85	2.538	0.4090	58	-1.15	009.09	0
009	Fiber, Neutral Detergent (%)	0265	22.15	0.1000	24.85	2.538	0.4090	58	-1.06	009.09	0
009	Fiber, Neutral Detergent (%)	0598	22.15	0.2600	24.85	2.538	0.4090	58	-1.06	009.09	0
009	Fiber, Neutral Detergent (%)	2193	22.18	0.8700	24.85	2.538	0.4090	58	-1.05	009.09	0
009	Fiber, Neutral Detergent (%)	0358	22.19	1.190	24.85	2.538	0.4090	58	-1.05	009.09	0
009	Fiber, Neutral Detergent (%)	0278	22.45	0.1000	24.85	2.538	0.4090	58	-0.94	009.09	0
009	Fiber, Neutral Detergent (%)	0263	22.59	0.1336	24.85	2.538	0.4090	58	-0.89	009.09	0
009	Fiber, Neutral Detergent (%)	0357	22.85	0.3000	24.85	2.538	0.4090	58	-0.79	009.09	0
009	Fiber, Neutral Detergent (%)	2196	22.92	0.0000	24.85	2.538	0.4090	58	-0.76	009.09	0
009	Fiber, Neutral Detergent (%)	0407	23.21	1.017	24.85	2.538	0.4090	58	-0.64	009.09	0
009	Fiber, Neutral Detergent (%)	0910	23.25	0.5000	24.85	2.538	0.4090	58	-0.63	009.07	0
009	Fiber, Neutral Detergent (%)	0948	23.83	0.5300	24.85	2.538	0.4090	58	-0.40	009.09	0
009	Fiber, Neutral Detergent (%)	2190	23.95	1.450	24.85	2.538	0.4090	58	-0.36	009.09	0
009	Fiber, Neutral Detergent (%)	0510	23.95	0.1000	24.85	2.538	0.4090	58	-0.35	009.09	0
009	Fiber, Neutral Detergent (%)	0904	24.08	0.3800	24.85	2.538	0.4090	58	-0.30	009.09	0
009	Fiber, Neutral Detergent (%)	0918	24.14	0.4261	24.85	2.538	0.4090	58	-0.28	009.09	0
009	Fiber, Neutral Detergent (%)	0504	24.36	0.3900	24.85	2.538	0.4090	58	-0.19	009.04	0
009	Fiber, Neutral Detergent (%)	0148	24.58	0.8000	24.85	2.538	0.4090	58	-0.11	009.09	0
009	Fiber, Neutral Detergent (%)	0646	24.75	0.7400	24.85	2.538	0.4090	58	-0.04	009.09	0
009	Fiber, Neutral Detergent (%)	0536	24.82	0.0400	24.85	2.538	0.4090	58	-0.01	009.09	0
009	Fiber, Neutral Detergent (%)	0354	24.86	0.4400	24.85	2.538	0.4090	58	0.01	009.09	0
009	Fiber, Neutral Detergent (%)	0889	25.25	0.3100	24.85	2.538	0.4090	58	0.16	009.09	0
009	Fiber, Neutral Detergent (%)	0226	25.30	0.1800	24.85	2.538	0.4090	58	0.18	009.07	0
009	Fiber, Neutral Detergent (%)	0619	25.40	0.2000	24.85	2.538	0.4090	58	0.22	009.09	0
009	Fiber, Neutral Detergent (%)	0001	25.47	0.4800	24.85	2.538	0.4090	58	0.25	009.09	0
009	Fiber, Neutral Detergent (%)	0353	25.57	0.2100	24.85	2.538	0.4090	58	0.28	009.07	0
009	Fiber, Neutral Detergent (%)	0638	25.70	0.2000	24.85	2.538	0.4090	58	0.34	009.09	0
009	Fiber, Neutral Detergent (%)	0848	25.72	0.0000	24.85	2.538	0.4090	58	0.34	009.09	0
009	Fiber, Neutral Detergent (%)	0885	25.78	0.0406	24.85	2.538	0.4090	58	0.37	009.07	0
009	Fiber, Neutral Detergent (%)	0590	26.00	0.2000	24.85	2.538	0.4090	58	0.45	009.07	0
009	Fiber, Neutral Detergent (%)	0653	26.09	0.5300	24.85	2.538	0.4090	58	0.49	009.09	0
009	Fiber, Neutral Detergent (%)	0345	26.09	0.2200	24.85	2.538	0.4090	58	0.49	009.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
009	Fiber, Neutral Detergent (%)	2192	26.25	1.150	24.85	2.538	0.4090	58	0.55	009.07	0
009	Fiber, Neutral Detergent (%)	0037	26.35	0.0200	24.85	2.538	0.4090	58	0.59	009.09	0
009	Fiber, Neutral Detergent (%)	0187	26.38	0.0100	24.85	2.538	0.4090	58	0.60	009.07	0
009	Fiber, Neutral Detergent (%)	0083	26.48	0.8900	24.85	2.538	0.4090	58	0.64	009.09	0
009	Fiber, Neutral Detergent (%)	0098	26.60	0.2000	24.85	2.538	0.4090	58	0.69	009.07	0
009	Fiber, Neutral Detergent (%)	0609	26.64	1.120	24.85	2.538	0.4090	58	0.71	009.07	0
009	Fiber, Neutral Detergent (%)	0066	26.68	0.3900	24.85	2.538	0.4090	58	0.72	009.09	0
009	Fiber, Neutral Detergent (%)	0941	27.15	0.1000	24.85	2.538	0.4090	58	0.91	009.99	0
009	Fiber, Neutral Detergent (%)	0889	27.18	1.190	24.85	2.538	0.4090	58	0.92	009.99	0
009	Fiber, Neutral Detergent (%)	0675	27.33	0.6100	24.85	2.538	0.4090	58	0.98	009.07	0
009	Fiber, Neutral Detergent (%)	0693	27.54	0.1480	24.85	2.538	0.4090	58	1.06	009.09	0
009	Fiber, Neutral Detergent (%)	0309	27.60	0.4284	24.85	2.538	0.4090	58	1.08	009.07	0
009	Fiber, Neutral Detergent (%)	0720	27.67	0.4000	24.85	2.538	0.4090	58	1.11	009.99	0
009	Fiber, Neutral Detergent (%)	0726	27.79	0.2200	24.85	2.538	0.4090	58	1.16	009.09	0
009	Fiber, Neutral Detergent (%)	0190	27.91	0.1800	24.85	2.538	0.4090	58	1.21	009.09	0
009	Fiber, Neutral Detergent (%)	0674	28.05	0.0000	24.85	2.538	0.4090	58	1.26	009.04	0
009	Fiber, Neutral Detergent (%)	0610	28.15	0.5000	24.85	2.538	0.4090	58	1.30	009.07	0
009	Fiber, Neutral Detergent (%)	0689	28.15	0.1000	24.85	2.538	0.4090	58	1.30	009.07	0
009	Fiber, Neutral Detergent (%)	2129	33.77	0.0200	24.85	2.538	0.4090	58	3.52	009.09	0
009	Fiber, Neutral Detergent (%)	0045	29.30	3.200	24.85	2.538	0.4090	58	1.75	009.09	1
010	Moisture (%)	0889	6.110	0.0200	7.549	1.030	0.1667	21	-1.40	010.11	0
010	Moisture (%)	2129	6.380	0.7600	7.549	1.030	0.1667	21	-1.14	010.99	0
010	Moisture (%)	0853	6.450	0.1000	7.549	1.030	0.1667	21	-1.07	010.99	0
010	Moisture (%)	0843	6.670	0.7000	7.549	1.030	0.1667	21	-0.85	010.03	0
010	Moisture (%)	2089	6.945	0.1700	7.549	1.030	0.1667	21	-0.59	010.99	0
010	Moisture (%)	0164	6.995	0.0900	7.549	1.030	0.1667	21	-0.54	010.03	0
010	Moisture (%)	2215	7.190	0.0400	7.549	1.030	0.1667	21	-0.35	010.99	0
010	Moisture (%)	0939	7.200	0.0000	7.549	1.030	0.1667	21	-0.34	010.99	0
010	Moisture (%)	0948	7.200	0.1000	7.549	1.030	0.1667	21	-0.34	010.99	0
010	Moisture (%)	0726	7.235	0.0300	7.549	1.030	0.1667	21	-0.31	010.99	0
010	Moisture (%)	0952	7.235	0.0300	7.549	1.030	0.1667	21	-0.31	010.99	0
010	Moisture (%)	0884	7.250	0.1000	7.549	1.030	0.1667	21	-0.29	010.99	0
010	Moisture (%)	2150	7.565	0.7500	7.549	1.030	0.1667	21	0.02	010.99	0
010	Moisture (%)	0921	7.750	0.0000	7.549	1.030	0.1667	21	0.19	010.99	0
010	Moisture (%)	0190	7.965	0.0300	7.549	1.030	0.1667	21	0.40	010.99	0
010	Moisture (%)	2076	8.125	0.0700	7.549	1.030	0.1667	21	0.56	010.99	0
010	Moisture (%)	2178	8.145	0.0700	7.549	1.030	0.1667	21	0.58	010.99	0
010	Moisture (%)	0152	8.845	0.0700	7.549	1.030	0.1667	21	1.26	010.11	0
010	Moisture (%)	0720	9.275	0.0100	7.549	1.030	0.1667	21	1.68	010.11	0
010	Moisture (%)	0405	9.710	0.0000	7.549	1.030	0.1667	21	2.10	010.99	0
010	Moisture (%)	0852	12.23	0.3600	7.549	1.030	0.1667	21	4.55	010.11	0
010	Moisture (%)	0692	6.050	1.100	7.549	1.030	0.1667	21	-1.46	010.99	1

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
011	Loss on Drying (%)	2198	6.530	0.1400	8.550	0.4366	0.1239	72	-4.63	011.01	0
011	Loss on Drying (%)	0175	6.700	0.2000	8.550	0.4366	0.1239	72	-4.24	011.01	0
011	Loss on Drying (%)	0853	6.800	0.0000	8.550	0.4366	0.1239	72	-4.01	011.01	0
011	Loss on Drying (%)	0880	6.975	0.2000	8.550	0.4366	0.1239	72	-3.61	011.01	0
011	Loss on Drying (%)	2031	7.167	0.1046	8.550	0.4366	0.1239	72	-3.17	011.01	0
011	Loss on Drying (%)	2073	7.635	0.0100	8.550	0.4366	0.1239	72	-2.10	011.01	0
011	Loss on Drying (%)	2199	7.650	0.1000	8.550	0.4366	0.1239	72	-2.06	011.01	0
011	Loss on Drying (%)	0598	7.795	0.4700	8.550	0.4366	0.1239	72	-1.73	011.01	0
011	Loss on Drying (%)	0643	7.840	0.4200	8.550	0.4366	0.1239	72	-1.63	011.01	0
011	Loss on Drying (%)	0226	7.955	0.0100	8.550	0.4366	0.1239	72	-1.36	011.01	0
011	Loss on Drying (%)	0511	7.970	0.0600	8.550	0.4366	0.1239	72	-1.33	011.01	0
011	Loss on Drying (%)	0874	8.045	0.2500	8.550	0.4366	0.1239	72	-1.16	011.01	0
011	Loss on Drying (%)	0536	8.075	0.2300	8.550	0.4366	0.1239	72	-1.09	011.01	0
011	Loss on Drying (%)	0541	8.140	0.3000	8.550	0.4366	0.1239	72	-0.94	011.01	0
011	Loss on Drying (%)	2200	8.200	0.4000	8.550	0.4366	0.1239	72	-0.80	011.01	0
011	Loss on Drying (%)	0100	8.240	0.5400	8.550	0.4366	0.1239	72	-0.71	011.01	0
011	Loss on Drying (%)	0953	8.280	0.3000	8.550	0.4366	0.1239	72	-0.62	011.01	0
011	Loss on Drying (%)	2181	8.315	0.1300	8.550	0.4366	0.1239	72	-0.54	011.01	0
011	Loss on Drying (%)	2193	8.320	0.1800	8.550	0.4366	0.1239	72	-0.53	011.01	0
011	Loss on Drying (%)	0164	8.335	0.0500	8.550	0.4366	0.1239	72	-0.49	011.01	0
011	Loss on Drying (%)	2223	8.355	0.0300	8.550	0.4366	0.1239	72	-0.45	011.02	0
011	Loss on Drying (%)	0202	8.391	0.0520	8.550	0.4366	0.1239	72	-0.36	011.01	0
011	Loss on Drying (%)	0857	8.400	0.0000	8.550	0.4366	0.1239	72	-0.34	011.99	0
011	Loss on Drying (%)	0968	8.425	0.0780	8.550	0.4366	0.1239	72	-0.29	011.01	0
011	Loss on Drying (%)	2075	8.445	0.1300	8.550	0.4366	0.1239	72	-0.24	011.02	0
011	Loss on Drying (%)	0723	8.450	0.0000	8.550	0.4366	0.1239	72	-0.23	011.01	0
011	Loss on Drying (%)	0653	8.465	0.0300	8.550	0.4366	0.1239	72	-0.19	011.01	0
011	Loss on Drying (%)	0425	8.500	0.0000	8.550	0.4366	0.1239	72	-0.11	011.02	0
011	Loss on Drying (%)	0539	8.510	0.1200	8.550	0.4366	0.1239	72	-0.09	011.01	0
011	Loss on Drying (%)	0407	8.530	0.0110	8.550	0.4366	0.1239	72	-0.05	011.01	0
011	Loss on Drying (%)	0098	8.540	0.1000	8.550	0.4366	0.1239	72	-0.02	011.01	0
011	Loss on Drying (%)	0354	8.540	0.0200	8.550	0.4366	0.1239	72	-0.02	011.01	0
011	Loss on Drying (%)	0682	8.540	0.0000	8.550	0.4366	0.1239	72	-0.02	011.01	0
011	Loss on Drying (%)	0881	8.572	0.2080	8.550	0.4366	0.1239	72	0.05	011.01	0
011	Loss on Drying (%)	2188	8.580	0.0600	8.550	0.4366	0.1239	72	0.07	011.01	0
011	Loss on Drying (%)	0309	8.585	0.1900	8.550	0.4366	0.1239	72	0.08	011.01	0
011	Loss on Drying (%)	0144	8.600	0.0400	8.550	0.4366	0.1239	72	0.11	011.01	0
011	Loss on Drying (%)	0955	8.615	0.1900	8.550	0.4366	0.1239	72	0.15	011.01	0
011	Loss on Drying (%)	0263	8.620	0.0400	8.550	0.4366	0.1239	72	0.16	011.01	0
011	Loss on Drying (%)	0651	8.633	0.0400	8.550	0.4366	0.1239	72	0.19	011.01	0
011	Loss on Drying (%)	0859	8.640	0.0120	8.550	0.4366	0.1239	72	0.21	011.01	0
011	Loss on Drying (%)	0520	8.650	0.1000	8.550	0.4366	0.1239	72	0.23	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
011	Loss on Drying (%)	0229	8.675	0.0300	8.550	0.4366	0.1239	72	0.29	011.01	0
011	Loss on Drying (%)	0171	8.680	0.0200	8.550	0.4366	0.1239	72	0.30	011.01	0
011	Loss on Drying (%)	0574	8.680	0.0400	8.550	0.4366	0.1239	72	0.30	011.01	0
011	Loss on Drying (%)	0646	8.690	0.0200	8.550	0.4366	0.1239	72	0.32	011.01	0
011	Loss on Drying (%)	2196	8.700	0.0000	8.550	0.4366	0.1239	72	0.34	011.01	0
011	Loss on Drying (%)	2022	8.725	0.0300	8.550	0.4366	0.1239	72	0.40	011.01	0
011	Loss on Drying (%)	0033	8.730	0.0800	8.550	0.4366	0.1239	72	0.41	011.01	0
011	Loss on Drying (%)	0294	8.750	0.1000	8.550	0.4366	0.1239	72	0.46	011.01	0
011	Loss on Drying (%)	0510	8.750	0.1000	8.550	0.4366	0.1239	72	0.46	011.01	0
011	Loss on Drying (%)	0242	8.760	0.0400	8.550	0.4366	0.1239	72	0.48	011.01	0
011	Loss on Drying (%)	0852	8.775	0.1700	8.550	0.4366	0.1239	72	0.52	011.99	0
011	Loss on Drying (%)	2006	8.780	0.3400	8.550	0.4366	0.1239	72	0.53	011.01	0
011	Loss on Drying (%)	0596	8.785	0.0500	8.550	0.4366	0.1239	72	0.54	011.01	0
011	Loss on Drying (%)	0870	8.826	0.2122	8.550	0.4366	0.1239	72	0.63	011.01	0
011	Loss on Drying (%)	0047	8.900	0.4000	8.550	0.4366	0.1239	72	0.80	011.01	0
011	Loss on Drying (%)	0589	8.900	0.3400	8.550	0.4366	0.1239	72	0.80	011.01	0
011	Loss on Drying (%)	0674	8.900	0.0000	8.550	0.4366	0.1239	72	0.80	011.01	0
011	Loss on Drying (%)	0553	8.905	0.0300	8.550	0.4366	0.1239	72	0.81	011.01	0
011	Loss on Drying (%)	0861	8.905	0.0500	8.550	0.4366	0.1239	72	0.81	011.01	0
011	Loss on Drying (%)	0618	8.965	0.0100	8.550	0.4366	0.1239	72	0.95	011.01	0
011	Loss on Drying (%)	0123	9.035	0.2500	8.550	0.4366	0.1239	72	1.11	011.01	0
011	Loss on Drying (%)	0205	9.065	0.1700	8.550	0.4366	0.1239	72	1.18	011.01	0
011	Loss on Drying (%)	0843	9.080	0.1000	8.550	0.4366	0.1239	72	1.21	011.02	0
011	Loss on Drying (%)	0848	9.085	0.0100	8.550	0.4366	0.1239	72	1.23	011.01	0
011	Loss on Drying (%)	0004	9.105	0.0100	8.550	0.4366	0.1239	72	1.27	011.01	0
011	Loss on Drying (%)	0685	9.105	0.3700	8.550	0.4366	0.1239	72	1.27	011.01	0
011	Loss on Drying (%)	0066	9.245	0.2100	8.550	0.4366	0.1239	72	1.59	011.01	0
011	Loss on Drying (%)	0265	9.300	0.0000	8.550	0.4366	0.1239	72	1.72	011.01	0
011	Loss on Drying (%)	0358	9.305	0.1900	8.550	0.4366	0.1239	72	1.73	011.01	0
011	Loss on Drying (%)	0119	9.480	0.0000	8.550	0.4366	0.1239	72	2.13	011.01	0
011	Loss on Drying (%)	0942	8.080	0.8400	8.550	0.4366	0.1239	72	-1.08	011.02	1
011	Loss on Drying (%)	0675	8.130	0.8800	8.550	0.4366	0.1239	72	-0.96	011.01	1
011	Loss on Drying (%)	0108	8.715	0.9900	8.550	0.4366	0.1239	72	0.38	011.01	1
012	Starch (%)	0904	7.920	0.4600	12.57	1.392	0.2173	40	-3.34	012.00	0
012	Starch (%)	0510	8.450	0.1000	12.57	1.392	0.2173	40	-2.96	012.04	0
012	Starch (%)	0889	8.985	0.1700	12.57	1.392	0.2173	40	-2.57	012.11	0
012	Starch (%)	0720	10.31	0.0700	12.57	1.392	0.2173	40	-1.63	012.11	0
012	Starch (%)	0278	10.80	0.4000	12.57	1.392	0.2173	40	-1.27	012.04	0
012	Starch (%)	0045	11.10	0.6000	12.57	1.392	0.2173	40	-1.05	012.01	0
012	Starch (%)	0160	11.12	0.4900	12.57	1.392	0.2173	40	-1.04	012.01	0
012	Starch (%)	0098	11.30	0.6000	12.57	1.392	0.2173	40	-0.91	012.03	0
012	Starch (%)	0148	11.33	0.1000	12.57	1.392	0.2173	40	-0.89	012.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
012	Starch (%)	0353	11.59	0.0100	12.57	1.392	0.2173	40	-0.71	012.04	0
012	Starch (%)	0164	11.60	0.0000	12.57	1.392	0.2173	40	-0.70	012.01	0
012	Starch (%)	0870	11.81	0.5872	12.57	1.392	0.2173	40	-0.54	012.01	0
012	Starch (%)	0407	11.90	0.1500	12.57	1.392	0.2173	40	-0.48	012.03	0
012	Starch (%)	0956	11.95	0.3000	12.57	1.392	0.2173	40	-0.44	012.01	0
012	Starch (%)	0676	11.99	0.3000	12.57	1.392	0.2173	40	-0.42	012.01	0
012	Starch (%)	2188	12.15	0.2600	12.57	1.392	0.2173	40	-0.30	012.00	0
012	Starch (%)	0164	12.55	0.1000	12.57	1.392	0.2173	40	-0.01	012.00	0
012	Starch (%)	0693	12.57	0.3740	12.57	1.392	0.2173	40	0.00	012.01	0
012	Starch (%)	2193	12.85	0.0900	12.57	1.392	0.2173	40	0.20	012.00	0
012	Starch (%)	0939	12.86	0.2700	12.57	1.392	0.2173	40	0.21	012.99	0
012	Starch (%)	0674	12.98	0.0000	12.57	1.392	0.2173	40	0.30	012.04	0
012	Starch (%)	2196	12.99	0.0000	12.57	1.392	0.2173	40	0.30	012.00	0
012	Starch (%)	0941	13.10	0.1100	12.57	1.392	0.2173	40	0.38	012.00	0
012	Starch (%)	0955	13.15	0.3000	12.57	1.392	0.2173	40	0.42	012.03	0
012	Starch (%)	2009	13.16	0.0093	12.57	1.392	0.2173	40	0.43	012.00	0
012	Starch (%)	0354	13.18	0.2700	12.57	1.392	0.2173	40	0.44	012.00	0
012	Starch (%)	2181	13.21	0.1100	12.57	1.392	0.2173	40	0.46	012.00	0
012	Starch (%)	0610	13.25	0.3000	12.57	1.392	0.2173	40	0.49	012.00	0
012	Starch (%)	0683	13.26	0.0000	12.57	1.392	0.2173	40	0.50	012.00	0
012	Starch (%)	2006	13.27	0.0200	12.57	1.392	0.2173	40	0.50	012.00	0
012	Starch (%)	0619	13.30	0.2000	12.57	1.392	0.2173	40	0.53	012.00	0
012	Starch (%)	0889	13.32	0.1700	12.57	1.392	0.2173	40	0.54	012.03	0
012	Starch (%)	0910	13.40	0.6000	12.57	1.392	0.2173	40	0.60	012.00	0
012	Starch (%)	2089	13.53	0.0400	12.57	1.392	0.2173	40	0.69	012.01	0
012	Starch (%)	0878	13.75	0.1000	12.57	1.392	0.2173	40	0.85	012.00	0
012	Starch (%)	0948	13.89	0.0200	12.57	1.392	0.2173	40	0.95	012.04	0
012	Starch (%)	0265	14.75	0.7000	12.57	1.392	0.2173	40	1.57	012.01	0
012	Starch (%)	0852	15.95	0.1700	12.57	1.392	0.2173	40	2.43	012.11	0
012	Starch (%)	2190	21.25	0.1000	12.57	1.392	0.2173	40	6.24	012.99	0
012	Starch (%)	0848	23.03	0.0400	12.57	1.392	0.2173	40	7.52	012.03	0
012	Starch (%)	0918	8.662	1.416	12.57	1.392	0.2173	40	-2.81	012.01	1
012	Starch (%)	2192	14.62	1.640	12.57	1.392	0.2173	40	1.47	012.00	1
012	Starch (%)	0609	0.0000	0.0000	12.57	1.392	0.2173	40		012.00	4
013	Fat, Acid Pretreat (%)	0643	0.0381	0.0039	3.232	1.194	0.1263	39	-2.68	013.02	0
013	Fat, Acid Pretreat (%)	0353	0.9200	0.1600	3.232	1.194	0.1263	39	-1.94	013.10	0
013	Fat, Acid Pretreat (%)	0638	1.750	0.0000	3.232	1.194	0.1263	39	-1.24	013.10	0
013	Fat, Acid Pretreat (%)	0083	1.975	0.1300	3.232	1.194	0.1263	39	-1.05	013.00	0
013	Fat, Acid Pretreat (%)	0689	2.000	0.0000	3.232	1.194	0.1263	39	-1.03	013.00	0
013	Fat, Acid Pretreat (%)	2196	2.050	0.0000	3.232	1.194	0.1263	39	-0.99	013.10	0
013	Fat, Acid Pretreat (%)	2006	2.080	0.0600	3.232	1.194	0.1263	39	-0.96	013.10	0
013	Fat, Acid Pretreat (%)	0242	2.220	0.0800	3.232	1.194	0.1263	39	-0.85	013.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
013	Fat, Acid Pretreat (%)	0297	2.295	0.0300	3.232	1.194	0.1263	39	-0.78	013.00	0
013	Fat, Acid Pretreat (%)	0948	2.295	0.0100	3.232	1.194	0.1263	39	-0.78	013.00	0
013	Fat, Acid Pretreat (%)	0874	2.335	0.0700	3.232	1.194	0.1263	39	-0.75	013.02	0
013	Fat, Acid Pretreat (%)	0884	2.450	0.1000	3.232	1.194	0.1263	39	-0.65	013.00	0
013	Fat, Acid Pretreat (%)	0610	2.530	0.1600	3.232	1.194	0.1263	39	-0.59	013.10	0
013	Fat, Acid Pretreat (%)	0618	2.553	0.1399	3.232	1.194	0.1263	39	-0.57	013.08	0
013	Fat, Acid Pretreat (%)	2215	2.660	0.0600	3.232	1.194	0.1263	39	-0.48	013.10	0
013	Fat, Acid Pretreat (%)	0910	2.765	0.1500	3.232	1.194	0.1263	39	-0.39	013.00	0
013	Fat, Acid Pretreat (%)	0853	2.865	0.0900	3.232	1.194	0.1263	39	-0.31	013.02	0
013	Fat, Acid Pretreat (%)	0676	2.970	0.1400	3.232	1.194	0.1263	39	-0.22	013.02	0
013	Fat, Acid Pretreat (%)	0354	3.200	0.2400	3.232	1.194	0.1263	39	-0.03	013.02	0
013	Fat, Acid Pretreat (%)	2073	3.250	0.0600	3.232	1.194	0.1263	39	0.02	013.00	0
013	Fat, Acid Pretreat (%)	0164	3.370	0.1200	3.232	1.194	0.1263	39	0.12	013.02	0
013	Fat, Acid Pretreat (%)	0148	3.455	0.0100	3.232	1.194	0.1263	39	0.19	013.02	0
013	Fat, Acid Pretreat (%)	0675	3.470	0.2200	3.232	1.194	0.1263	39	0.20	013.02	0
013	Fat, Acid Pretreat (%)	0229	3.685	0.3500	3.232	1.194	0.1263	39	0.38	013.00	0
013	Fat, Acid Pretreat (%)	0541	3.690	0.2000	3.232	1.194	0.1263	39	0.38	013.00	0
013	Fat, Acid Pretreat (%)	0870	3.742	0.0001	3.232	1.194	0.1263	39	0.43	013.02	0
013	Fat, Acid Pretreat (%)	0682	3.800	0.0000	3.232	1.194	0.1263	39	0.48	013.02	0
013	Fat, Acid Pretreat (%)	0100	3.845	0.5700	3.232	1.194	0.1263	39	0.51	013.02	0
013	Fat, Acid Pretreat (%)	0309	3.914	0.3487	3.232	1.194	0.1263	39	0.57	013.00	0
013	Fat, Acid Pretreat (%)	0553	4.065	0.3900	3.232	1.194	0.1263	39	0.70	013.02	0
013	Fat, Acid Pretreat (%)	0202	4.278	0.0950	3.232	1.194	0.1263	39	0.88	013.00	0
013	Fat, Acid Pretreat (%)	0407	4.287	0.1285	3.232	1.194	0.1263	39	0.88	013.13	0
013	Fat, Acid Pretreat (%)	0171	4.290	0.0600	3.232	1.194	0.1263	39	0.89	013.02	0
013	Fat, Acid Pretreat (%)	2150	4.390	0.0000	3.232	1.194	0.1263	39	0.97	013.00	0
013	Fat, Acid Pretreat (%)	0843	4.685	0.0100	3.232	1.194	0.1263	39	1.22	013.13	0
013	Fat, Acid Pretreat (%)	0861	4.880	0.1000	3.232	1.194	0.1263	39	1.38	013.02	0
013	Fat, Acid Pretreat (%)	2076	5.213	0.0422	3.232	1.194	0.1263	39	1.66	013.00	0
013	Fat, Acid Pretreat (%)	0618	6.409	0.0270	3.232	1.194	0.1263	39	2.66	013.00	0
013	Fat, Acid Pretreat (%)	0265	8.195	0.5700	3.232	1.194	0.1263	39	4.16	013.13	0
013	Fat, Acid Pretreat (%)	0504	3.260	0.7400	3.232	1.194	0.1263	39	0.02	013.00	1
015	Aluminum (mg / kg (ppm))	0164	358.5	53.00	600.0	63.45	19.79	12	-3.81	015.43	0
015	Aluminum (mg / kg (ppm))	0353	439.0	4.000	600.0	63.45	19.79	12	-2.54	015.43	0
015	Aluminum (mg / kg (ppm))	0049	554.3	36.12	600.0	63.45	19.79	12	-0.72	015.41	0
015	Aluminum (mg / kg (ppm))	0918	575.6	28.50	600.0	63.45	19.79	12	-0.39	015.43	0
015	Aluminum (mg / kg (ppm))	0169	597.0	66.00	600.0	63.45	19.79	12	-0.05	015.43	0
015	Aluminum (mg / kg (ppm))	0186	597.0	4.000	600.0	63.45	19.79	12	-0.05	015.52	0
015	Aluminum (mg / kg (ppm))	0171	625.2	5.900	600.0	63.45	19.79	12	0.40	015.41	0
015	Aluminum (mg / kg (ppm))	0510	625.5	1.000	600.0	63.45	19.79	12	0.40	015.43	0
015	Aluminum (mg / kg (ppm))	0425	635.1	8.000	600.0	63.45	19.79	12	0.55	015.43	0
015	Aluminum (mg / kg (ppm))	0520	652.0	28.00	600.0	63.45	19.79	12	0.82	015.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
015	Aluminum (mg / kg (ppm))	0345	658.0	2.000	600.0	63.45	19.79	12	0.91	015.43	0
015	Aluminum (mg / kg (ppm))	0560	671.1	1.000	600.0	63.45	19.79	12	1.12	015.42	0
017	Boron (mg / kg (ppm))	0032	0.6350	0.0100	25.09	13.55	1.115	16	-1.80	017.53	0
017	Boron (mg / kg (ppm))	0560	14.35	3.300	25.09	13.55	1.115	16	-0.79	017.42	0
017	Boron (mg / kg (ppm))	2053	17.10	1.200	25.09	13.55	1.115	16	-0.59	017.42	0
017	Boron (mg / kg (ppm))	0358	17.42	0.5500	25.09	13.55	1.115	16	-0.57	017.41	0
017	Boron (mg / kg (ppm))	0265	17.45	1.700	25.09	13.55	1.115	16	-0.56	017.42	0
017	Boron (mg / kg (ppm))	0693	18.67	0.5860	25.09	13.55	1.115	16	-0.47	017.42	0
017	Boron (mg / kg (ppm))	0049	19.00	0.2500	25.09	13.55	1.115	16	-0.45	017.41	0
017	Boron (mg / kg (ppm))	0171	19.15	0.7400	25.09	13.55	1.115	16	-0.44	017.41	0
017	Boron (mg / kg (ppm))	0345	19.27	0.5300	25.09	13.55	1.115	16	-0.43	017.43	0
017	Boron (mg / kg (ppm))	0918	27.81	0.3000	25.09	13.55	1.115	16	0.20	017.43	0
017	Boron (mg / kg (ppm))	0510	29.00	0.0000	25.09	13.55	1.115	16	0.29	017.43	0
017	Boron (mg / kg (ppm))	0553	30.45	2.100	25.09	13.55	1.115	16	0.40	017.41	0
017	Boron (mg / kg (ppm))	0226	32.15	0.3400	25.09	13.55	1.115	16	0.52	017.41	0
017	Boron (mg / kg (ppm))	0353	44.00	2.000	25.09	13.55	1.115	16	1.40	017.43	0
017	Boron (mg / kg (ppm))	0229	51.56	2.640	25.09	13.55	1.115	16	1.95	017.41	0
017	Boron (mg / kg (ppm))	0294	63.30	1.600	25.09	13.55	1.115	16	2.82	017.42	0
017	Boron (mg / kg (ppm))	0045	15.23	11.55	25.09	13.55	1.115	16	-0.73	017.42	1
019	Calcium (%)	2200	2.300	0.2000	4.087	0.2066	0.0793	136	-8.65	019.08	0
019	Calcium (%)	0629	3.230	0.1400	4.087	0.2066	0.0793	136	-4.15	019.08	0
019	Calcium (%)	0205	3.440	0.0200	4.087	0.2066	0.0793	136	-3.13	019.42	0
019	Calcium (%)	2129	3.445	0.0500	4.087	0.2066	0.0793	136	-3.11	019.42	0
019	Calcium (%)	0609	3.555	0.0300	4.087	0.2066	0.0793	136	-2.57	019.32	0
019	Calcium (%)	0144	3.575	0.1900	4.087	0.2066	0.0793	136	-2.48	019.41	0
019	Calcium (%)	0018	3.580	0.0200	4.087	0.2066	0.0793	136	-2.45	019.31	0
019	Calcium (%)	0405	3.610	0.0000	4.087	0.2066	0.0793	136	-2.31	019.41	0
019	Calcium (%)	2055	3.625	0.1551	4.087	0.2066	0.0793	136	-2.23	019.53	0
019	Calcium (%)	0016	3.650	0.1400	4.087	0.2066	0.0793	136	-2.11	019.52	0
019	Calcium (%)	2126	3.705	0.0300	4.087	0.2066	0.0793	136	-1.85	019.31	0
019	Calcium (%)	0868	3.740	0.0400	4.087	0.2066	0.0793	136	-1.68	019.31	0
019	Calcium (%)	2006	3.774	0.0510	4.087	0.2066	0.0793	136	-1.52	019.09	0
019	Calcium (%)	0511	3.780	0.3800	4.087	0.2066	0.0793	136	-1.48	019.41	0
019	Calcium (%)	0675	3.810	0.0200	4.087	0.2066	0.0793	136	-1.34	019.43	0
019	Calcium (%)	0045	3.820	0.2600	4.087	0.2066	0.0793	136	-1.29	019.42	0
019	Calcium (%)	0036	3.852	0.0647	4.087	0.2066	0.0793	136	-1.14	019.43	0
019	Calcium (%)	0345	3.855	0.0100	4.087	0.2066	0.0793	136	-1.12	019.43	0
019	Calcium (%)	2006	3.860	0.0170	4.087	0.2066	0.0793	136	-1.10	019.00	0
019	Calcium (%)	0098	3.860	0.0000	4.087	0.2066	0.0793	136	-1.10	019.43	0
019	Calcium (%)	0553	3.865	0.2100	4.087	0.2066	0.0793	136	-1.07	019.41	0
019	Calcium (%)	2022	3.880	0.0000	4.087	0.2066	0.0793	136	-1.00	019.31	0
019	Calcium (%)	2145	3.880	0.0086	4.087	0.2066	0.0793	136	-1.00	019.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	0263	3.881	0.0490	4.087	0.2066	0.0793	136	-0.99	019.42	0
019	Calcium (%)	0357	3.887	0.0121	4.087	0.2066	0.0793	136	-0.97	019.42	0
019	Calcium (%)	0123	3.915	0.0100	4.087	0.2066	0.0793	136	-0.83	019.41	0
019	Calcium (%)	0619	3.915	0.1300	4.087	0.2066	0.0793	136	-0.83	019.41	0
019	Calcium (%)	0202	3.917	0.0440	4.087	0.2066	0.0793	136	-0.82	019.43	0
019	Calcium (%)	0939	3.925	0.4500	4.087	0.2066	0.0793	136	-0.78	019.53	0
019	Calcium (%)	0278	3.930	0.0600	4.087	0.2066	0.0793	136	-0.76	019.42	0
019	Calcium (%)	0035	3.935	0.0300	4.087	0.2066	0.0793	136	-0.73	019.43	0
019	Calcium (%)	0242	3.950	0.0200	4.087	0.2066	0.0793	136	-0.66	019.99	0
019	Calcium (%)	0300	3.955	0.2900	4.087	0.2066	0.0793	136	-0.64	019.33	0
019	Calcium (%)	0098	3.965	0.0300	4.087	0.2066	0.0793	136	-0.59	019.41	0
019	Calcium (%)	0009	3.968	0.3103	4.087	0.2066	0.0793	136	-0.58	019.42	0
019	Calcium (%)	2031	3.975	0.0506	4.087	0.2066	0.0793	136	-0.54	019.08	0
019	Calcium (%)	0169	3.975	0.0100	4.087	0.2066	0.0793	136	-0.54	019.32	0
019	Calcium (%)	0520	3.980	0.0800	4.087	0.2066	0.0793	136	-0.52	019.41	0
019	Calcium (%)	0674	3.980	0.0000	4.087	0.2066	0.0793	136	-0.52	019.31	0
019	Calcium (%)	0171	3.985	0.0100	4.087	0.2066	0.0793	136	-0.49	019.41	0
019	Calcium (%)	0948	3.986	0.0560	4.087	0.2066	0.0793	136	-0.49	019.31	0
019	Calcium (%)	2181	3.990	0.0200	4.087	0.2066	0.0793	136	-0.47	019.00	0
019	Calcium (%)	0010	3.990	0.0800	4.087	0.2066	0.0793	136	-0.47	019.43	0
019	Calcium (%)	2109	3.992	0.1220	4.087	0.2066	0.0793	136	-0.46	019.00	0
019	Calcium (%)	0175	4.000	0.0000	4.087	0.2066	0.0793	136	-0.42	019.31	0
019	Calcium (%)	2195	4.000	0.0000	4.087	0.2066	0.0793	136	-0.42	019.31	0
019	Calcium (%)	0510	4.005	0.0100	4.087	0.2066	0.0793	136	-0.40	019.43	0
019	Calcium (%)	0590	4.005	0.0300	4.087	0.2066	0.0793	136	-0.40	019.99	0
019	Calcium (%)	0265	4.010	0.0600	4.087	0.2066	0.0793	136	-0.37	019.42	0
019	Calcium (%)	0874	4.015	0.0100	4.087	0.2066	0.0793	136	-0.35	019.31	0
019	Calcium (%)	0848	4.015	0.0100	4.087	0.2066	0.0793	136	-0.35	019.41	0
019	Calcium (%)	0019	4.030	0.2800	4.087	0.2066	0.0793	136	-0.27	019.41	0
019	Calcium (%)	0720	4.035	0.0700	4.087	0.2066	0.0793	136	-0.25	019.41	0
019	Calcium (%)	2062	4.045	0.0300	4.087	0.2066	0.0793	136	-0.20	019.00	0
019	Calcium (%)	0042	4.045	0.2500	4.087	0.2066	0.0793	136	-0.20	019.42	0
019	Calcium (%)	2192	4.050	0.0600	4.087	0.2066	0.0793	136	-0.18	019.43	0
019	Calcium (%)	0164	4.055	0.1700	4.087	0.2066	0.0793	136	-0.15	019.41	0
019	Calcium (%)	0033	4.055	0.0100	4.087	0.2066	0.0793	136	-0.15	019.43	0
019	Calcium (%)	0893	4.060	0.0000	4.087	0.2066	0.0793	136	-0.13	019.00	0
019	Calcium (%)	0142	4.064	0.0731	4.087	0.2066	0.0793	136	-0.11	019.31	0
019	Calcium (%)	0560	4.064	0.0367	4.087	0.2066	0.0793	136	-0.11	019.42	0
019	Calcium (%)	0425	4.065	0.0300	4.087	0.2066	0.0793	136	-0.11	019.43	0
019	Calcium (%)	0610	4.070	0.0200	4.087	0.2066	0.0793	136	-0.08	019.43	0
019	Calcium (%)	2089	4.075	0.0700	4.087	0.2066	0.0793	136	-0.06	019.43	0
019	Calcium (%)	0035	4.080	0.0600	4.087	0.2066	0.0793	136	-0.03	019.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	0003	4.085	0.3100	4.087	0.2066	0.0793	136	-0.01	019.41	0
019	Calcium (%)	0687	4.090	0.0600	4.087	0.2066	0.0793	136	0.02	019.31	0
019	Calcium (%)	0964	4.095	0.1119	4.087	0.2066	0.0793	136	0.04	019.43	0
019	Calcium (%)	0027	4.104	0.0260	4.087	0.2066	0.0793	136	0.08	019.43	0
019	Calcium (%)	0083	4.110	0.0400	4.087	0.2066	0.0793	136	0.11	019.43	0
019	Calcium (%)	2215	4.115	0.0700	4.087	0.2066	0.0793	136	0.14	019.99	0
019	Calcium (%)	0885	4.116	0.0324	4.087	0.2066	0.0793	136	0.14	019.08	0
019	Calcium (%)	0881	4.120	0.0440	4.087	0.2066	0.0793	136	0.16	019.00	0
019	Calcium (%)	2196	4.120	0.0000	4.087	0.2066	0.0793	136	0.16	019.08	0
019	Calcium (%)	0968	4.120	0.0100	4.087	0.2066	0.0793	136	0.16	019.43	0
019	Calcium (%)	2022	4.125	0.0100	4.087	0.2066	0.0793	136	0.18	019.00	0
019	Calcium (%)	0148	4.125	0.0100	4.087	0.2066	0.0793	136	0.18	019.41	0
019	Calcium (%)	0693	4.132	0.0660	4.087	0.2066	0.0793	136	0.22	019.42	0
019	Calcium (%)	2012	4.135	0.0100	4.087	0.2066	0.0793	136	0.23	019.42	0
019	Calcium (%)	0651	4.144	0.0070	4.087	0.2066	0.0793	136	0.27	019.00	0
019	Calcium (%)	0186	4.145	0.1100	4.087	0.2066	0.0793	136	0.28	019.52	0
019	Calcium (%)	0849	4.147	0.0130	4.087	0.2066	0.0793	136	0.29	019.08	0
019	Calcium (%)	2196	4.150	0.0000	4.087	0.2066	0.0793	136	0.31	019.31	0
019	Calcium (%)	0878	4.155	0.0500	4.087	0.2066	0.0793	136	0.33	019.00	0
019	Calcium (%)	0074	4.155	0.0100	4.087	0.2066	0.0793	136	0.33	019.41	0
019	Calcium (%)	0294	4.155	0.0100	4.087	0.2066	0.0793	136	0.33	019.42	0
019	Calcium (%)	0723	4.160	0.0000	4.087	0.2066	0.0793	136	0.35	019.31	0
019	Calcium (%)	0910	4.160	0.1800	4.087	0.2066	0.0793	136	0.35	019.41	0
019	Calcium (%)	0861	4.160	0.0200	4.087	0.2066	0.0793	136	0.35	019.43	0
019	Calcium (%)	0689	4.165	0.0500	4.087	0.2066	0.0793	136	0.38	019.08	0
019	Calcium (%)	0870	4.165	0.2000	4.087	0.2066	0.0793	136	0.38	019.42	0
019	Calcium (%)	0047	4.170	0.0400	4.087	0.2066	0.0793	136	0.40	019.52	0
019	Calcium (%)	0187	4.175	0.0100	4.087	0.2066	0.0793	136	0.43	019.42	0
019	Calcium (%)	2009	4.189	0.0025	4.087	0.2066	0.0793	136	0.49	019.08	0
019	Calcium (%)	0505	4.190	0.0600	4.087	0.2066	0.0793	136	0.50	019.43	0
019	Calcium (%)	0089	4.190	0.0000	4.087	0.2066	0.0793	136	0.50	019.43	0
019	Calcium (%)	0683	4.195	0.0300	4.087	0.2066	0.0793	136	0.52	019.00	0
019	Calcium (%)	2188	4.200	0.1000	4.087	0.2066	0.0793	136	0.55	019.08	0
019	Calcium (%)	0001	4.201	0.0410	4.087	0.2066	0.0793	136	0.55	019.31	0
019	Calcium (%)	0152	4.205	0.1300	4.087	0.2066	0.0793	136	0.57	019.31	0
019	Calcium (%)	0897	4.215	0.0100	4.087	0.2066	0.0793	136	0.62	019.00	0
019	Calcium (%)	0692	4.215	0.0300	4.087	0.2066	0.0793	136	0.62	019.42	0
019	Calcium (%)	2190	4.219	0.0743	4.087	0.2066	0.0793	136	0.64	019.08	0
019	Calcium (%)	0108	4.225	0.3500	4.087	0.2066	0.0793	136	0.67	019.31	0
019	Calcium (%)	0572	4.225	0.0300	4.087	0.2066	0.0793	136	0.67	019.53	0
019	Calcium (%)	0006	4.227	0.0550	4.087	0.2066	0.0793	136	0.68	019.43	0
019	Calcium (%)	0354	4.230	0.0400	4.087	0.2066	0.0793	136	0.69	019.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
019	Calcium (%)	0553	4.235	0.1900	4.087	0.2066	0.0793	136	0.72	019.53	0
019	Calcium (%)	0229	4.235	0.1700	4.087	0.2066	0.0793	136	0.72	019.41	0
019	Calcium (%)	0598	4.238	0.0320	4.087	0.2066	0.0793	136	0.73	019.41	0
019	Calcium (%)	0049	4.240	0.0400	4.087	0.2066	0.0793	136	0.74	019.41	0
019	Calcium (%)	0590	4.245	0.0500	4.087	0.2066	0.0793	136	0.77	019.08	0
019	Calcium (%)	0682	4.260	0.0000	4.087	0.2066	0.0793	136	0.84	019.43	0
019	Calcium (%)	2193	4.270	0.0200	4.087	0.2066	0.0793	136	0.89	019.08	0
019	Calcium (%)	0186	4.270	0.3600	4.087	0.2066	0.0793	136	0.89	019.42	0
019	Calcium (%)	2198	4.275	0.0500	4.087	0.2066	0.0793	136	0.91	019.31	0
019	Calcium (%)	0685	4.300	0.0400	4.087	0.2066	0.0793	136	1.03	019.42	0
019	Calcium (%)	0656	4.310	0.0200	4.087	0.2066	0.0793	136	1.08	019.31	0
019	Calcium (%)	0726	4.316	0.0000	4.087	0.2066	0.0793	136	1.11	019.42	0
019	Calcium (%)	0653	4.322	0.0010	4.087	0.2066	0.0793	136	1.14	019.41	0
019	Calcium (%)	0612	4.340	0.0800	4.087	0.2066	0.0793	136	1.23	019.33	0
019	Calcium (%)	0878	4.350	0.1000	4.087	0.2066	0.0793	136	1.27	019.41	0
019	Calcium (%)	2053	4.350	0.1600	4.087	0.2066	0.0793	136	1.27	019.42	0
019	Calcium (%)	0037	4.365	0.0300	4.087	0.2066	0.0793	136	1.35	019.42	0
019	Calcium (%)	0852	4.400	0.2000	4.087	0.2066	0.0793	136	1.52	019.08	0
019	Calcium (%)	0100	4.440	0.3000	4.087	0.2066	0.0793	136	1.71	019.99	0
019	Calcium (%)	0407	4.451	0.2280	4.087	0.2066	0.0793	136	1.76	019.41	0
019	Calcium (%)	0676	4.455	0.0100	4.087	0.2066	0.0793	136	1.78	019.99	0
019	Calcium (%)	0226	4.465	0.1900	4.087	0.2066	0.0793	136	1.83	019.41	0
019	Calcium (%)	0190	4.480	0.0800	4.087	0.2066	0.0793	136	1.90	019.42	0
019	Calcium (%)	0358	4.510	0.0200	4.087	0.2066	0.0793	136	2.05	019.41	0
019	Calcium (%)	0036	4.512	0.0470	4.087	0.2066	0.0793	136	2.06	019.03	0
019	Calcium (%)	0028	4.520	0.3200	4.087	0.2066	0.0793	136	2.10	019.43	0
019	Calcium (%)	0353	4.540	0.0600	4.087	0.2066	0.0793	136	2.19	019.43	0
019	Calcium (%)	2128	5.166	0.2900	4.087	0.2066	0.0793	136	5.22	019.02	0
019	Calcium (%)	0032	5.430	0.0600	4.087	0.2066	0.0793	136	6.50	019.43	0
019	Calcium (%)	0536	3.625	0.5900	4.087	0.2066	0.0793	136	-2.24	019.31	1
019	Calcium (%)	0884	3.840	0.6000	4.087	0.2066	0.0793	136	-1.19	019.31	1
019	Calcium (%)	0918	4.521	0.5430	4.087	0.2066	0.0793	136	2.10	019.43	1
019	Calcium (%)	0154	4.573	0.6644	4.087	0.2066	0.0793	136	2.35	019.52	1
019	Calcium (%)	2223	1.215	0.0300	4.087	0.2066	0.0793	136	-13.90	019.00	2
021	Cobalt (mg / kg (ppm))	0955	35.50	4.800	46.68	5.163	1.434	27	-2.17	021.41	0
021	Cobalt (mg / kg (ppm))	0345	37.66	0.3200	46.68	5.163	1.434	27	-1.75	021.43	0
021	Cobalt (mg / kg (ppm))	0619	40.20	0.2000	46.68	5.163	1.434	27	-1.26	021.41	0
021	Cobalt (mg / kg (ppm))	0047	40.45	1.030	46.68	5.163	1.434	27	-1.21	021.52	0
021	Cobalt (mg / kg (ppm))	0164	42.00	6.000	46.68	5.163	1.434	27	-0.91	021.31	0
021	Cobalt (mg / kg (ppm))	0016	42.35	0.7000	46.68	5.163	1.434	27	-0.84	021.52	0
021	Cobalt (mg / kg (ppm))	0169	43.95	0.5000	46.68	5.163	1.434	27	-0.53	021.43	0
021	Cobalt (mg / kg (ppm))	0171	44.06	0.3900	46.68	5.163	1.434	27	-0.51	021.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
021	Cobalt (mg / kg (ppm))	0610	45.00	2.000	46.68	5.163	1.434	27	-0.33	021.34	0
021	Cobalt (mg / kg (ppm))	0693	45.12	1.230	46.68	5.163	1.434	27	-0.30	021.42	0
021	Cobalt (mg / kg (ppm))	0510	45.34	0.2200	46.68	5.163	1.434	27	-0.26	021.43	0
021	Cobalt (mg / kg (ppm))	2055	45.56	4.198	46.68	5.163	1.434	27	-0.22	021.53	0
021	Cobalt (mg / kg (ppm))	0045	45.95	1.700	46.68	5.163	1.434	27	-0.14	021.42	0
021	Cobalt (mg / kg (ppm))	0186	46.00	2.000	46.68	5.163	1.434	27	-0.13	021.42	0
021	Cobalt (mg / kg (ppm))	0560	46.60	0.4000	46.68	5.163	1.434	27	-0.02	021.42	0
021	Cobalt (mg / kg (ppm))	0148	46.74	0.2900	46.68	5.163	1.434	27	0.01	021.41	0
021	Cobalt (mg / kg (ppm))	0910	47.23	5.020	46.68	5.163	1.434	27	0.11	021.52	0
021	Cobalt (mg / kg (ppm))	0186	49.00	0.0000	46.68	5.163	1.434	27	0.45	021.52	0
021	Cobalt (mg / kg (ppm))	0964	49.88	0.7432	46.68	5.163	1.434	27	0.62	021.43	0
021	Cobalt (mg / kg (ppm))	0848	50.14	0.1000	46.68	5.163	1.434	27	0.67	021.41	0
021	Cobalt (mg / kg (ppm))	0027	50.26	0.9750	46.68	5.163	1.434	27	0.69	021.43	0
021	Cobalt (mg / kg (ppm))	0918	51.49	0.5800	46.68	5.163	1.434	27	0.93	021.53	0
021	Cobalt (mg / kg (ppm))	0098	51.57	2.560	46.68	5.163	1.434	27	0.95	021.53	0
021	Cobalt (mg / kg (ppm))	0013	52.15	0.5000	46.68	5.163	1.434	27	1.06	021.43	0
021	Cobalt (mg / kg (ppm))	0553	52.70	0.4000	46.68	5.163	1.434	27	1.17	021.53	0
021	Cobalt (mg / kg (ppm))	0939	54.65	0.6600	46.68	5.163	1.434	27	1.54	021.53	0
021	Cobalt (mg / kg (ppm))	0032	58.90	1.200	46.68	5.163	1.434	27	2.37	021.53	0
022	Copper (mg / kg (ppm))	2192	565.0	103.4	866.2	58.78	17.40	98	-5.12	022.43	0
022	Copper (mg / kg (ppm))	0405	629.3	0.0000	866.2	58.78	17.40	98	-4.03	022.41	0
022	Copper (mg / kg (ppm))	0598	695.8	37.80	866.2	58.78	17.40	98	-2.90	022.41	0
022	Copper (mg / kg (ppm))	0553	751.5	21.00	866.2	58.78	17.40	98	-1.95	022.53	0
022	Copper (mg / kg (ppm))	0553	763.0	20.00	866.2	58.78	17.40	98	-1.76	022.41	0
022	Copper (mg / kg (ppm))	2055	765.0	38.00	866.2	58.78	17.40	98	-1.72	022.53	0
022	Copper (mg / kg (ppm))	0345	766.5	9.000	866.2	58.78	17.40	98	-1.70	022.43	0
022	Copper (mg / kg (ppm))	0629	768.0	0.0000	866.2	58.78	17.40	98	-1.67	022.31	0
022	Copper (mg / kg (ppm))	0049	785.7	22.14	866.2	58.78	17.40	98	-1.37	022.41	0
022	Copper (mg / kg (ppm))	0190	786.6	2.750	866.2	58.78	17.40	98	-1.35	022.42	0
022	Copper (mg / kg (ppm))	2145	801.0	4.770	866.2	58.78	17.40	98	-1.11	022.41	0
022	Copper (mg / kg (ppm))	2128	801.7	83.32	866.2	58.78	17.40	98	-1.10	022.32	0
022	Copper (mg / kg (ppm))	0849	803.5	19.00	866.2	58.78	17.40	98	-1.07	022.31	0
022	Copper (mg / kg (ppm))	0511	808.0	22.00	866.2	58.78	17.40	98	-0.99	022.41	0
022	Copper (mg / kg (ppm))	0520	808.0	40.00	866.2	58.78	17.40	98	-0.99	022.41	0
022	Copper (mg / kg (ppm))	0202	808.8	4.145	866.2	58.78	17.40	98	-0.98	022.43	0
022	Copper (mg / kg (ppm))	0407	809.7	15.12	866.2	58.78	17.40	98	-0.96	022.41	0
022	Copper (mg / kg (ppm))	0019	810.1	10.80	866.2	58.78	17.40	98	-0.95	022.41	0
022	Copper (mg / kg (ppm))	0035	812.0	48.00	866.2	58.78	17.40	98	-0.92	022.43	0
022	Copper (mg / kg (ppm))	0689	815.7	24.00	866.2	58.78	17.40	98	-0.86	022.31	0
022	Copper (mg / kg (ppm))	2089	822.3	4.470	866.2	58.78	17.40	98	-0.75	022.43	0
022	Copper (mg / kg (ppm))	2144	825.0	8.940	866.2	58.78	17.40	98	-0.70	022.31	0
022	Copper (mg / kg (ppm))	0910	825.5	59.00	866.2	58.78	17.40	98	-0.69	022.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
022	Copper (mg / kg (ppm))	0263	825.6	2.830	866.2	58.78	17.40	98	-0.69	022.42	0
022	Copper (mg / kg (ppm))	0653	827.6	12.80	866.2	58.78	17.40	98	-0.66	022.41	0
022	Copper (mg / kg (ppm))	0510	830.0	2.000	866.2	58.78	17.40	98	-0.62	022.43	0
022	Copper (mg / kg (ppm))	0242	830.0	2.000	866.2	58.78	17.40	98	-0.62	022.99	0
022	Copper (mg / kg (ppm))	0003	832.5	13.00	866.2	58.78	17.40	98	-0.57	022.41	0
022	Copper (mg / kg (ppm))	0848	833.2	4.290	866.2	58.78	17.40	98	-0.56	022.41	0
022	Copper (mg / kg (ppm))	0045	835.5	41.00	866.2	58.78	17.40	98	-0.52	022.42	0
022	Copper (mg / kg (ppm))	0009	837.5	23.00	866.2	58.78	17.40	98	-0.49	022.42	0
022	Copper (mg / kg (ppm))	0171	838.5	15.60	866.2	58.78	17.40	98	-0.47	022.41	0
022	Copper (mg / kg (ppm))	0964	845.8	17.91	866.2	58.78	17.40	98	-0.35	022.43	0
022	Copper (mg / kg (ppm))	0037	846.5	25.00	866.2	58.78	17.40	98	-0.34	022.42	0
022	Copper (mg / kg (ppm))	0619	847.5	11.00	866.2	58.78	17.40	98	-0.32	022.41	0
022	Copper (mg / kg (ppm))	0868	848.0	8.330	866.2	58.78	17.40	98	-0.31	022.31	0
022	Copper (mg / kg (ppm))	0572	849.0	40.00	866.2	58.78	17.40	98	-0.29	022.53	0
022	Copper (mg / kg (ppm))	2022	850.8	0.8120	866.2	58.78	17.40	98	-0.26	022.31	0
022	Copper (mg / kg (ppm))	0278	852.8	29.90	866.2	58.78	17.40	98	-0.23	022.42	0
022	Copper (mg / kg (ppm))	2062	854.0	6.000	866.2	58.78	17.40	98	-0.21	022.31	0
022	Copper (mg / kg (ppm))	0001	854.9	17.50	866.2	58.78	17.40	98	-0.19	022.31	0
022	Copper (mg / kg (ppm))	0425	856.0	6.100	866.2	58.78	17.40	98	-0.17	022.43	0
022	Copper (mg / kg (ppm))	0042	856.5	3.000	866.2	58.78	17.40	98	-0.17	022.42	0
022	Copper (mg / kg (ppm))	0357	857.7	4.673	866.2	58.78	17.40	98	-0.14	022.42	0
022	Copper (mg / kg (ppm))	0536	858.6	2.500	866.2	58.78	17.40	98	-0.13	022.31	0
022	Copper (mg / kg (ppm))	0027	860.3	9.593	866.2	58.78	17.40	98	-0.10	022.43	0
022	Copper (mg / kg (ppm))	0169	862.5	29.00	866.2	58.78	17.40	98	-0.06	022.43	0
022	Copper (mg / kg (ppm))	0968	864.0	2.000	866.2	58.78	17.40	98	-0.04	022.43	0
022	Copper (mg / kg (ppm))	0560	864.2	6.800	866.2	58.78	17.40	98	-0.03	022.42	0
022	Copper (mg / kg (ppm))	0010	866.5	5.700	866.2	58.78	17.40	98	0.00	022.33	0
022	Copper (mg / kg (ppm))	0682	867.7	0.0000	866.2	58.78	17.40	98	0.03	022.43	0
022	Copper (mg / kg (ppm))	0354	870.8	1.140	866.2	58.78	17.40	98	0.08	022.31	0
022	Copper (mg / kg (ppm))	0098	871.9	12.80	866.2	58.78	17.40	98	0.10	022.43	0
022	Copper (mg / kg (ppm))	2195	873.0	4.570	866.2	58.78	17.40	98	0.12	022.31	0
022	Copper (mg / kg (ppm))	0870	873.3	11.40	866.2	58.78	17.40	98	0.12	022.42	0
022	Copper (mg / kg (ppm))	0035	874.0	0.0000	866.2	58.78	17.40	98	0.13	022.42	0
022	Copper (mg / kg (ppm))	0033	874.5	33.00	866.2	58.78	17.40	98	0.14	022.43	0
022	Copper (mg / kg (ppm))	0028	875.7	36.46	866.2	58.78	17.40	98	0.16	022.43	0
022	Copper (mg / kg (ppm))	0148	876.1	17.62	866.2	58.78	17.40	98	0.17	022.41	0
022	Copper (mg / kg (ppm))	0726	878.5	2.900	866.2	58.78	17.40	98	0.21	022.42	0
022	Copper (mg / kg (ppm))	2146	879.5	21.00	866.2	58.78	17.40	98	0.23	022.51	0
022	Copper (mg / kg (ppm))	0939	883.1	30.18	866.2	58.78	17.40	98	0.29	022.53	0
022	Copper (mg / kg (ppm))	0187	883.3	4.320	866.2	58.78	17.40	98	0.29	022.42	0
022	Copper (mg / kg (ppm))	0693	884.7	14.11	866.2	58.78	17.40	98	0.31	022.42	0
022	Copper (mg / kg (ppm))	0074	885.5	23.00	866.2	58.78	17.40	98	0.33	022.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
022	Copper (mg / kg (ppm))	0505	887.3	31.10	866.2	58.78	17.40	98	0.36	022.43	0
022	Copper (mg / kg (ppm))	0100	889.5	7.000	866.2	58.78	17.40	98	0.40	022.99	0
022	Copper (mg / kg (ppm))	0948	890.3	0.5160	866.2	58.78	17.40	98	0.41	022.33	0
022	Copper (mg / kg (ppm))	2196	891.6	0.0000	866.2	58.78	17.40	98	0.43	022.31	0
022	Copper (mg / kg (ppm))	0205	894.0	4.000	866.2	58.78	17.40	98	0.47	022.42	0
022	Copper (mg / kg (ppm))	0175	895.0	10.00	866.2	58.78	17.40	98	0.49	022.31	0
022	Copper (mg / kg (ppm))	0918	897.3	45.98	866.2	58.78	17.40	98	0.53	022.43	0
022	Copper (mg / kg (ppm))	0186	901.0	14.00	866.2	58.78	17.40	98	0.59	022.52	0
022	Copper (mg / kg (ppm))	0098	907.5	23.60	866.2	58.78	17.40	98	0.70	022.41	0
022	Copper (mg / kg (ppm))	0083	908.5	5.000	866.2	58.78	17.40	98	0.72	022.43	0
022	Copper (mg / kg (ppm))	0353	908.5	9.000	866.2	58.78	17.40	98	0.72	022.43	0
022	Copper (mg / kg (ppm))	0861	910.5	35.00	866.2	58.78	17.40	98	0.75	022.43	0
022	Copper (mg / kg (ppm))	0723	912.0	0.0000	866.2	58.78	17.40	98	0.78	022.31	0
022	Copper (mg / kg (ppm))	0229	914.2	16.70	866.2	58.78	17.40	98	0.82	022.41	0
022	Copper (mg / kg (ppm))	0610	914.5	23.00	866.2	58.78	17.40	98	0.82	022.43	0
022	Copper (mg / kg (ppm))	0226	917.9	42.86	866.2	58.78	17.40	98	0.88	022.41	0
022	Copper (mg / kg (ppm))	0164	919.5	3.000	866.2	58.78	17.40	98	0.91	022.41	0
022	Copper (mg / kg (ppm))	0656	921.2	27.71	866.2	58.78	17.40	98	0.94	022.31	0
022	Copper (mg / kg (ppm))	0186	927.0	22.00	866.2	58.78	17.40	98	1.03	022.42	0
022	Copper (mg / kg (ppm))	0016	934.5	43.00	866.2	58.78	17.40	98	1.16	022.52	0
022	Copper (mg / kg (ppm))	0265	939.5	17.00	866.2	58.78	17.40	98	1.25	022.42	0
022	Copper (mg / kg (ppm))	0685	940.8	1.790	866.2	58.78	17.40	98	1.27	022.42	0
022	Copper (mg / kg (ppm))	0294	945.8	10.80	866.2	58.78	17.40	98	1.35	022.42	0
022	Copper (mg / kg (ppm))	0720	947.2	32.15	866.2	58.78	17.40	98	1.38	022.41	0
022	Copper (mg / kg (ppm))	0590	954.0	8.000	866.2	58.78	17.40	98	1.49	022.99	0
022	Copper (mg / kg (ppm))	0878	957.0	22.00	866.2	58.78	17.40	98	1.54	022.41	0
022	Copper (mg / kg (ppm))	2188	963.7	6.110	866.2	58.78	17.40	98	1.66	022.31	0
022	Copper (mg / kg (ppm))	0358	972.8	18.26	866.2	58.78	17.40	98	1.81	022.41	0
022	Copper (mg / kg (ppm))	0032	980.5	11.00	866.2	58.78	17.40	98	1.94	022.43	0
022	Copper (mg / kg (ppm))	2129	1,046	1.000	866.2	58.78	17.40	98	3.05	022.99	0
022	Copper (mg / kg (ppm))	0047	1,053	31.70	866.2	58.78	17.40	98	3.18	022.52	0
022	Copper (mg / kg (ppm))	2012	1,258	19.28	866.2	58.78	17.40	98	6.66	022.42	0
022	Copper (mg / kg (ppm))	0675	1,364	10.32	866.2	58.78	17.40	98	8.47	022.43	0
022	Copper (mg / kg (ppm))	0692	1,176	428.0	866.2	58.78	17.40	98	5.27	022.42	1
023	Fluorine (mg / kg (ppm))	0910	159.5	10.00				1		023.01	0
024	Iodine (mg / kg (ppm))	0171	7.950	0.5000				2		024.03	0
024	Iodine (mg / kg (ppm))	0186	42.00	0.0000				2		024.52	0
025	Iron (mg / kg (ppm))	0006	1.059	0.0020	10,593	1,332	271.2	86	-7.95	025.43	0
025	Iron (mg / kg (ppm))	0964	1.085	0.0826	10,593	1,332	271.2	86	-7.95	025.43	0
025	Iron (mg / kg (ppm))	0619	1.175	0.0500	10,593	1,332	271.2	86	-7.95	025.41	0
025	Iron (mg / kg (ppm))	2195	969.7	70.32	10,593	1,332	271.2	86	-7.22	025.31	0
025	Iron (mg / kg (ppm))	0726	1,020	16.00	10,593	1,332	271.2	86	-7.19	025.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
025	Iron (mg / kg (ppm))	0226	1,813	141.1	10,593	1,332	271.2	86	-6.59	025.41	0
025	Iron (mg / kg (ppm))	0405	4,444	0.0000	10,593	1,332	271.2	86	-4.62	025.41	0
025	Iron (mg / kg (ppm))	0598	4,551	281.0	10,593	1,332	271.2	86	-4.54	025.41	0
025	Iron (mg / kg (ppm))	0720	5,467	748.0	10,593	1,332	271.2	86	-3.85	025.41	0
025	Iron (mg / kg (ppm))	0675	5,695	249.8	10,593	1,332	271.2	86	-3.68	025.43	0
025	Iron (mg / kg (ppm))	0037	7,176	272.0	10,593	1,332	271.2	86	-2.57	025.42	0
025	Iron (mg / kg (ppm))	0187	7,471	69.83	10,593	1,332	271.2	86	-2.34	025.42	0
025	Iron (mg / kg (ppm))	0948	8,721	101.6	10,593	1,332	271.2	86	-1.41	025.33	0
025	Iron (mg / kg (ppm))	0278	9,102	786.0	10,593	1,332	271.2	86	-1.12	025.42	0
025	Iron (mg / kg (ppm))	0692	9,200	200.0	10,593	1,332	271.2	86	-1.05	025.42	0
025	Iron (mg / kg (ppm))	0164	9,350	80.00	10,593	1,332	271.2	86	-0.93	025.41	0
025	Iron (mg / kg (ppm))	0345	9,355	70.00	10,593	1,332	271.2	86	-0.93	025.43	0
025	Iron (mg / kg (ppm))	0870	9,476	650.0	10,593	1,332	271.2	86	-0.84	025.42	0
025	Iron (mg / kg (ppm))	0016	9,495	170.0	10,593	1,332	271.2	86	-0.82	025.52	0
025	Iron (mg / kg (ppm))	0035	9,534	139.0	10,593	1,332	271.2	86	-0.80	025.43	0
025	Iron (mg / kg (ppm))	0003	9,574	324.0	10,593	1,332	271.2	86	-0.76	025.41	0
025	Iron (mg / kg (ppm))	0689	9,581	189.5	10,593	1,332	271.2	86	-0.76	025.31	0
025	Iron (mg / kg (ppm))	0553	9,789	70.00	10,593	1,332	271.2	86	-0.60	025.41	0
025	Iron (mg / kg (ppm))	0939	9,842	475.0	10,593	1,332	271.2	86	-0.56	025.53	0
025	Iron (mg / kg (ppm))	0047	9,998	385.0	10,593	1,332	271.2	86	-0.45	025.52	0
025	Iron (mg / kg (ppm))	0045	10,050	100.0	10,593	1,332	271.2	86	-0.41	025.42	0
025	Iron (mg / kg (ppm))	0190	10,163	77.00	10,593	1,332	271.2	86	-0.32	025.42	0
025	Iron (mg / kg (ppm))	0098	10,200	0.0000	10,593	1,332	271.2	86	-0.30	025.43	0
025	Iron (mg / kg (ppm))	2188	10,362	74.10	10,593	1,332	271.2	86	-0.17	025.31	0
025	Iron (mg / kg (ppm))	0505	10,385	70.00	10,593	1,332	271.2	86	-0.16	025.31	0
025	Iron (mg / kg (ppm))	2146	10,416	123.0	10,593	1,332	271.2	86	-0.13	025.51	0
025	Iron (mg / kg (ppm))	2012	10,438	1,601	10,593	1,332	271.2	86	-0.12	025.42	0
025	Iron (mg / kg (ppm))	2055	10,497	1,120	10,593	1,332	271.2	86	-0.07	025.53	0
025	Iron (mg / kg (ppm))	0358	10,566	434.9	10,593	1,332	271.2	86	-0.02	025.41	0
025	Iron (mg / kg (ppm))	0019	10,595	530.0	10,593	1,332	271.2	86	0.00	025.41	0
025	Iron (mg / kg (ppm))	0510	10,612	23.00	10,593	1,332	271.2	86	0.01	025.43	0
025	Iron (mg / kg (ppm))	0098	10,650	300.0	10,593	1,332	271.2	86	0.04	025.41	0
025	Iron (mg / kg (ppm))	0511	10,680	20.00	10,593	1,332	271.2	86	0.07	025.41	0
025	Iron (mg / kg (ppm))	2053	10,738	1,015	10,593	1,332	271.2	86	0.11	025.42	0
025	Iron (mg / kg (ppm))	0083	10,740	140.0	10,593	1,332	271.2	86	0.11	025.43	0
025	Iron (mg / kg (ppm))	0653	10,741	53.00	10,593	1,332	271.2	86	0.11	025.41	0
025	Iron (mg / kg (ppm))	0169	10,754	508.0	10,593	1,332	271.2	86	0.12	025.43	0
025	Iron (mg / kg (ppm))	0918	10,768	1,041	10,593	1,332	271.2	86	0.13	025.43	0
025	Iron (mg / kg (ppm))	0520	10,835	271.0	10,593	1,332	271.2	86	0.18	025.41	0
025	Iron (mg / kg (ppm))	0175	10,850	900.0	10,593	1,332	271.2	86	0.19	025.31	0
025	Iron (mg / kg (ppm))	0263	10,866	112.5	10,593	1,332	271.2	86	0.21	025.42	0
025	Iron (mg / kg (ppm))	0425	10,871	29.00	10,593	1,332	271.2	86	0.21	025.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
025	Iron (mg / kg (ppm))	0723	10,897	0.0000	10,593	1,332	271.2	86	0.23	025.31	0
025	Iron (mg / kg (ppm))	2022	10,939	16.90	10,593	1,332	271.2	86	0.26	025.31	0
025	Iron (mg / kg (ppm))	0042	10,955	870.0	10,593	1,332	271.2	86	0.27	025.42	0
025	Iron (mg / kg (ppm))	0868	11,019	1,163	10,593	1,332	271.2	86	0.32	025.31	0
025	Iron (mg / kg (ppm))	2062	11,036	14.40	10,593	1,332	271.2	86	0.33	025.31	0
025	Iron (mg / kg (ppm))	0560	11,048	670.2	10,593	1,332	271.2	86	0.34	025.42	0
025	Iron (mg / kg (ppm))	0049	11,115	135.3	10,593	1,332	271.2	86	0.39	025.41	0
025	Iron (mg / kg (ppm))	0148	11,126	158.8	10,593	1,332	271.2	86	0.40	025.41	0
025	Iron (mg / kg (ppm))	0035	11,145	190.0	10,593	1,332	271.2	86	0.41	025.42	0
025	Iron (mg / kg (ppm))	0874	11,150	100.0	10,593	1,332	271.2	86	0.42	025.31	0
025	Iron (mg / kg (ppm))	0265	11,150	100.0	10,593	1,332	271.2	86	0.42	025.42	0
025	Iron (mg / kg (ppm))	0033	11,150	500.0	10,593	1,332	271.2	86	0.42	025.43	0
025	Iron (mg / kg (ppm))	0878	11,152	26.00	10,593	1,332	271.2	86	0.42	025.41	0
025	Iron (mg / kg (ppm))	2144	11,174	0.0000	10,593	1,332	271.2	86	0.44	025.31	0
025	Iron (mg / kg (ppm))	0074	11,211	208.0	10,593	1,332	271.2	86	0.46	025.41	0
025	Iron (mg / kg (ppm))	0001	11,270	60.00	10,593	1,332	271.2	86	0.51	025.31	0
025	Iron (mg / kg (ppm))	0610	11,307	15.00	10,593	1,332	271.2	86	0.54	025.43	0
025	Iron (mg / kg (ppm))	0968	11,343	48.00	10,593	1,332	271.2	86	0.56	025.43	0
025	Iron (mg / kg (ppm))	0693	11,413	237.1	10,593	1,332	271.2	86	0.62	025.42	0
025	Iron (mg / kg (ppm))	0294	11,442	57.38	10,593	1,332	271.2	86	0.64	025.42	0
025	Iron (mg / kg (ppm))	0027	11,480	95.11	10,593	1,332	271.2	86	0.67	025.43	0
025	Iron (mg / kg (ppm))	2196	11,518	0.0000	10,593	1,332	271.2	86	0.69	025.31	0
025	Iron (mg / kg (ppm))	0407	11,573	639.2	10,593	1,332	271.2	86	0.74	025.41	0
025	Iron (mg / kg (ppm))	0354	11,649	48.09	10,593	1,332	271.2	86	0.79	025.31	0
025	Iron (mg / kg (ppm))	0553	11,750	100.0	10,593	1,332	271.2	86	0.87	025.53	0
025	Iron (mg / kg (ppm))	2145	11,767	32.00	10,593	1,332	271.2	86	0.88	025.41	0
025	Iron (mg / kg (ppm))	0229	11,797	51.00	10,593	1,332	271.2	86	0.90	025.41	0
025	Iron (mg / kg (ppm))	0656	11,818	57.06	10,593	1,332	271.2	86	0.92	025.31	0
025	Iron (mg / kg (ppm))	0682	11,849	0.0000	10,593	1,332	271.2	86	0.94	025.43	0
025	Iron (mg / kg (ppm))	2128	11,927	1,128	10,593	1,332	271.2	86	1.00	025.32	0
025	Iron (mg / kg (ppm))	0861	12,150	300.0	10,593	1,332	271.2	86	1.17	025.43	0
025	Iron (mg / kg (ppm))	0910	12,219	250.0	10,593	1,332	271.2	86	1.22	025.41	0
025	Iron (mg / kg (ppm))	2129	12,260	240.0	10,593	1,332	271.2	86	1.25	025.99	0
025	Iron (mg / kg (ppm))	0100	12,329	282.0	10,593	1,332	271.2	86	1.30	025.99	0
025	Iron (mg / kg (ppm))	0353	12,907	417.0	10,593	1,332	271.2	86	1.74	025.43	0
025	Iron (mg / kg (ppm))	0171	13,000	260.0	10,593	1,332	271.2	86	1.81	025.41	0
025	Iron (mg / kg (ppm))	0536	13,603	715.0	10,593	1,332	271.2	86	2.26	025.31	0
025	Iron (mg / kg (ppm))	0848	14,883	6.940	10,593	1,332	271.2	86	3.22	025.41	0
025	Iron (mg / kg (ppm))	0032	15,450	100.0	10,593	1,332	271.2	86	3.65	025.43	0
025	Iron (mg / kg (ppm))	2192	4,709	1,883	10,593	1,332	271.2	86	-4.42	025.43	1
025	Iron (mg / kg (ppm))	0028	15,846	2,798	10,593	1,332	271.2	86	3.94	025.43	1
027	Magnesium (%)	0353	0.2650	0.0300	0.4851	0.0351	0.0128	98	-6.27	027.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
027	Magnesium (%)	0405	0.4000	0.0000	0.4851	0.0351	0.0128	98	-2.43	027.41	0
027	Magnesium (%)	0089	0.4000	0.0000	0.4851	0.0351	0.0128	98	-2.43	027.43	0
027	Magnesium (%)	0596	0.4050	0.0300	0.4851	0.0351	0.0128	98	-2.28	027.31	0
027	Magnesium (%)	2129	0.4170	0.0000	0.4851	0.0351	0.0128	98	-1.94	027.42	0
027	Magnesium (%)	0629	0.4200	0.0000	0.4851	0.0351	0.0128	98	-1.86	027.31	0
027	Magnesium (%)	0510	0.4300	0.0000	0.4851	0.0351	0.0128	98	-1.57	027.43	0
027	Magnesium (%)	2022	0.4320	0.0040	0.4851	0.0351	0.0128	98	-1.51	027.31	0
027	Magnesium (%)	0169	0.4320	0.0380	0.4851	0.0351	0.0128	98	-1.51	027.32	0
027	Magnesium (%)	0035	0.4350	0.0100	0.4851	0.0351	0.0128	98	-1.43	027.43	0
027	Magnesium (%)	0598	0.4357	0.0109	0.4851	0.0351	0.0128	98	-1.41	027.41	0
027	Magnesium (%)	0345	0.4370	0.0080	0.4851	0.0351	0.0128	98	-1.37	027.43	0
027	Magnesium (%)	0049	0.4400	0.0200	0.4851	0.0351	0.0128	98	-1.29	027.41	0
027	Magnesium (%)	0202	0.4450	0.0160	0.4851	0.0351	0.0128	98	-1.14	027.43	0
027	Magnesium (%)	0553	0.4480	0.0180	0.4851	0.0351	0.0128	98	-1.06	027.41	0
027	Magnesium (%)	0407	0.4503	0.0043	0.4851	0.0351	0.0128	98	-0.99	027.41	0
027	Magnesium (%)	0918	0.4520	0.0360	0.4851	0.0351	0.0128	98	-0.94	027.43	0
027	Magnesium (%)	0870	0.4543	0.0180	0.4851	0.0351	0.0128	98	-0.88	027.42	0
027	Magnesium (%)	0425	0.4550	0.0100	0.4851	0.0351	0.0128	98	-0.86	027.43	0
027	Magnesium (%)	0045	0.4565	0.0090	0.4851	0.0351	0.0128	98	-0.81	027.42	0
027	Magnesium (%)	0939	0.4600	0.0400	0.4851	0.0351	0.0128	98	-0.72	027.53	0
027	Magnesium (%)	2145	0.4623	0.0030	0.4851	0.0351	0.0128	98	-0.65	027.41	0
027	Magnesium (%)	0027	0.4645	0.0050	0.4851	0.0351	0.0128	98	-0.59	027.43	0
027	Magnesium (%)	0536	0.4650	0.0100	0.4851	0.0351	0.0128	98	-0.57	027.31	0
027	Magnesium (%)	0689	0.4650	0.0100	0.4851	0.0351	0.0128	98	-0.57	027.31	0
027	Magnesium (%)	0003	0.4650	0.0100	0.4851	0.0351	0.0128	98	-0.57	027.41	0
027	Magnesium (%)	0520	0.4650	0.0300	0.4851	0.0351	0.0128	98	-0.57	027.41	0
027	Magnesium (%)	0968	0.4670	0.0020	0.4851	0.0351	0.0128	98	-0.52	027.43	0
027	Magnesium (%)	0016	0.4685	0.0150	0.4851	0.0351	0.0128	98	-0.47	027.52	0
027	Magnesium (%)	0693	0.4695	0.0070	0.4851	0.0351	0.0128	98	-0.44	027.42	0
027	Magnesium (%)	0098	0.4700	0.0000	0.4851	0.0351	0.0128	98	-0.43	027.43	0
027	Magnesium (%)	0505	0.4700	0.0000	0.4851	0.0351	0.0128	98	-0.43	027.43	0
027	Magnesium (%)	0242	0.4700	0.0000	0.4851	0.0351	0.0128	98	-0.43	027.99	0
027	Magnesium (%)	0861	0.4730	0.0160	0.4851	0.0351	0.0128	98	-0.34	027.43	0
027	Magnesium (%)	0010	0.4750	0.0100	0.4851	0.0351	0.0128	98	-0.29	027.33	0
027	Magnesium (%)	0148	0.4750	0.0100	0.4851	0.0351	0.0128	98	-0.29	027.41	0
027	Magnesium (%)	0511	0.4750	0.0100	0.4851	0.0351	0.0128	98	-0.29	027.41	0
027	Magnesium (%)	0035	0.4750	0.0100	0.4851	0.0351	0.0128	98	-0.29	027.42	0
027	Magnesium (%)	0263	0.4761	0.0034	0.4851	0.0351	0.0128	98	-0.26	027.42	0
027	Magnesium (%)	0009	0.4778	0.0307	0.4851	0.0351	0.0128	98	-0.21	027.42	0
027	Magnesium (%)	2144	0.4779	0.0020	0.4851	0.0351	0.0128	98	-0.21	027.31	0
027	Magnesium (%)	0294	0.4800	0.0200	0.4851	0.0351	0.0128	98	-0.15	027.42	0
027	Magnesium (%)	0033	0.4800	0.0000	0.4851	0.0351	0.0128	98	-0.15	027.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
027	Magnesium (%)	0868	0.4850	0.0100	0.4851	0.0351	0.0128	98	0.00	027.31	0
027	Magnesium (%)	0019	0.4850	0.0300	0.4851	0.0351	0.0128	98	0.00	027.41	0
027	Magnesium (%)	0692	0.4850	0.0100	0.4851	0.0351	0.0128	98	0.00	027.42	0
027	Magnesium (%)	0083	0.4850	0.0100	0.4851	0.0351	0.0128	98	0.00	027.43	0
027	Magnesium (%)	0142	0.4868	0.0015	0.4851	0.0351	0.0128	98	0.05	027.31	0
027	Magnesium (%)	0572	0.4880	0.0340	0.4851	0.0351	0.0128	98	0.08	027.53	0
027	Magnesium (%)	0560	0.4886	0.0024	0.4851	0.0351	0.0128	98	0.10	027.42	0
027	Magnesium (%)	0164	0.4895	0.0310	0.4851	0.0351	0.0128	98	0.13	027.41	0
027	Magnesium (%)	0357	0.4899	0.0013	0.4851	0.0351	0.0128	98	0.14	027.42	0
027	Magnesium (%)	0175	0.4900	0.0000	0.4851	0.0351	0.0128	98	0.14	027.31	0
027	Magnesium (%)	0910	0.4900	0.0400	0.4851	0.0351	0.0128	98	0.14	027.41	0
027	Magnesium (%)	0685	0.4900	0.0000	0.4851	0.0351	0.0128	98	0.14	027.42	0
027	Magnesium (%)	0047	0.4900	0.0200	0.4851	0.0351	0.0128	98	0.14	027.52	0
027	Magnesium (%)	0590	0.4900	0.0000	0.4851	0.0351	0.0128	98	0.14	027.99	0
027	Magnesium (%)	0619	0.4905	0.0030	0.4851	0.0351	0.0128	98	0.15	027.41	0
027	Magnesium (%)	0186	0.4910	0.0380	0.4851	0.0351	0.0128	98	0.17	027.42	0
027	Magnesium (%)	0874	0.4940	0.0080	0.4851	0.0351	0.0128	98	0.25	027.31	0
027	Magnesium (%)	0948	0.4950	0.0100	0.4851	0.0351	0.0128	98	0.28	027.33	0
027	Magnesium (%)	0098	0.4950	0.0100	0.4851	0.0351	0.0128	98	0.28	027.41	0
027	Magnesium (%)	0278	0.4950	0.0100	0.4851	0.0351	0.0128	98	0.28	027.42	0
027	Magnesium (%)	0074	0.4965	0.0070	0.4851	0.0351	0.0128	98	0.33	027.41	0
027	Magnesium (%)	0187	0.4965	0.0024	0.4851	0.0351	0.0128	98	0.33	027.42	0
027	Magnesium (%)	0186	0.4980	0.0100	0.4851	0.0351	0.0128	98	0.37	027.52	0
027	Magnesium (%)	0354	0.5000	0.0000	0.4851	0.0351	0.0128	98	0.42	027.31	0
027	Magnesium (%)	0609	0.5000	0.0200	0.4851	0.0351	0.0128	98	0.42	027.32	0
027	Magnesium (%)	0229	0.5000	0.0200	0.4851	0.0351	0.0128	98	0.42	027.41	0
027	Magnesium (%)	0878	0.5000	0.0200	0.4851	0.0351	0.0128	98	0.42	027.41	0
027	Magnesium (%)	2089	0.5000	0.0200	0.4851	0.0351	0.0128	98	0.42	027.43	0
027	Magnesium (%)	0006	0.5025	0.0010	0.4851	0.0351	0.0128	98	0.50	027.43	0
027	Magnesium (%)	0964	0.5026	0.0134	0.4851	0.0351	0.0128	98	0.50	027.43	0
027	Magnesium (%)	2146	0.5088	0.0037	0.4851	0.0351	0.0128	98	0.67	027.51	0
027	Magnesium (%)	0042	0.5090	0.0040	0.4851	0.0351	0.0128	98	0.68	027.42	0
027	Magnesium (%)	0610	0.5100	0.0000	0.4851	0.0351	0.0128	98	0.71	027.43	0
027	Magnesium (%)	0612	0.5110	0.0020	0.4851	0.0351	0.0128	98	0.74	027.33	0
027	Magnesium (%)	0553	0.5140	0.0320	0.4851	0.0351	0.0128	98	0.82	027.53	0
027	Magnesium (%)	0358	0.5150	0.0100	0.4851	0.0351	0.0128	98	0.85	027.41	0
027	Magnesium (%)	0190	0.5150	0.0100	0.4851	0.0351	0.0128	98	0.85	027.42	0
027	Magnesium (%)	2012	0.5150	0.0100	0.4851	0.0351	0.0128	98	0.85	027.42	0
027	Magnesium (%)	0726	0.5156	0.0004	0.4851	0.0351	0.0128	98	0.87	027.42	0
027	Magnesium (%)	0653	0.5205	0.0010	0.4851	0.0351	0.0128	98	1.01	027.41	0
027	Magnesium (%)	0001	0.5247	0.0403	0.4851	0.0351	0.0128	98	1.13	027.31	0
027	Magnesium (%)	0226	0.5250	0.0100	0.4851	0.0351	0.0128	98	1.14	027.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
027	Magnesium (%)	0675	0.5250	0.0100	0.4851	0.0351	0.0128	98	1.14	027.43	0
027	Magnesium (%)	0100	0.5250	0.0300	0.4851	0.0351	0.0128	98	1.14	027.99	0
027	Magnesium (%)	2196	0.5300	0.0000	0.4851	0.0351	0.0128	98	1.28	027.31	0
027	Magnesium (%)	0720	0.5300	0.0000	0.4851	0.0351	0.0128	98	1.28	027.41	0
027	Magnesium (%)	0028	0.5300	0.1200	0.4851	0.0351	0.0128	98	1.28	027.43	0
027	Magnesium (%)	0265	0.5350	0.0100	0.4851	0.0351	0.0128	98	1.42	027.42	0
027	Magnesium (%)	0682	0.5400	0.0000	0.4851	0.0351	0.0128	98	1.56	027.43	0
027	Magnesium (%)	0656	0.5450	0.0100	0.4851	0.0351	0.0128	98	1.71	027.31	0
027	Magnesium (%)	0171	0.5495	0.0050	0.4851	0.0351	0.0128	98	1.84	027.41	0
027	Magnesium (%)	0037	0.5500	0.0200	0.4851	0.0351	0.0128	98	1.85	027.42	0
027	Magnesium (%)	0154	0.5579	0.0296	0.4851	0.0351	0.0128	98	2.07	027.52	0
027	Magnesium (%)	2053	0.5750	0.0100	0.4851	0.0351	0.0128	98	2.56	027.42	0
027	Magnesium (%)	0032	0.6000	0.0000	0.4851	0.0351	0.0128	98	3.27	027.43	0
027	Magnesium (%)	2055	4,545	254.0	0.4851	0.0351	0.0128	98	129509.80	027.53	2
027	Magnesium (%)	0848	6,259	48.51	0.4851	0.0351	0.0128	98	178345.06	027.41	2
028	Manganese (mg / kg (ppm))	0720	131.3	3.840	276.6	22.01	6.875	94	-6.60	028.41	0
028	Manganese (mg / kg (ppm))	2062	136.2	0.7000	276.6	22.01	6.875	94	-6.38	028.31	0
028	Manganese (mg / kg (ppm))	0405	185.8	0.0000	276.6	22.01	6.875	94	-4.12	028.41	0
028	Manganese (mg / kg (ppm))	0598	222.6	7.600	276.6	22.01	6.875	94	-2.45	028.41	0
028	Manganese (mg / kg (ppm))	2192	223.4	32.72	276.6	22.01	6.875	94	-2.42	028.53	0
028	Manganese (mg / kg (ppm))	0049	223.9	3.460	276.6	22.01	6.875	94	-2.39	028.41	0
028	Manganese (mg / kg (ppm))	2055	227.5	6.800	276.6	22.01	6.875	94	-2.23	028.53	0
028	Manganese (mg / kg (ppm))	0037	247.0	10.00	276.6	22.01	6.875	94	-1.35	028.42	0
028	Manganese (mg / kg (ppm))	0629	247.5	5.000	276.6	22.01	6.875	94	-1.32	028.31	0
028	Manganese (mg / kg (ppm))	0553	248.5	25.00	276.6	22.01	6.875	94	-1.28	028.53	0
028	Manganese (mg / kg (ppm))	2195	251.7	1.410	276.6	22.01	6.875	94	-1.13	028.31	0
028	Manganese (mg / kg (ppm))	0511	252.0	20.00	276.6	22.01	6.875	94	-1.12	028.41	0
028	Manganese (mg / kg (ppm))	0520	252.5	13.00	276.6	22.01	6.875	94	-1.10	028.41	0
028	Manganese (mg / kg (ppm))	0536	253.1	2.880	276.6	22.01	6.875	94	-1.07	028.31	0
028	Manganese (mg / kg (ppm))	2129	256.0	2.000	276.6	22.01	6.875	94	-0.94	028.99	0
028	Manganese (mg / kg (ppm))	0553	257.5	3.000	276.6	22.01	6.875	94	-0.87	028.41	0
028	Manganese (mg / kg (ppm))	0278	257.5	21.00	276.6	22.01	6.875	94	-0.87	028.42	0
028	Manganese (mg / kg (ppm))	0407	258.1	27.72	276.6	22.01	6.875	94	-0.84	028.41	0
028	Manganese (mg / kg (ppm))	0003	258.5	1.000	276.6	22.01	6.875	94	-0.82	028.41	0
028	Manganese (mg / kg (ppm))	0689	258.9	5.600	276.6	22.01	6.875	94	-0.81	028.31	0
028	Manganese (mg / kg (ppm))	0870	258.9	13.60	276.6	22.01	6.875	94	-0.81	028.42	0
028	Manganese (mg / kg (ppm))	0190	259.6	0.1200	276.6	22.01	6.875	94	-0.78	028.42	0
028	Manganese (mg / kg (ppm))	2022	259.9	2.752	276.6	22.01	6.875	94	-0.76	028.31	0
028	Manganese (mg / kg (ppm))	0964	260.5	0.7820	276.6	22.01	6.875	94	-0.73	028.43	0
028	Manganese (mg / kg (ppm))	0175	261.0	2.000	276.6	22.01	6.875	94	-0.71	028.31	0
028	Manganese (mg / kg (ppm))	0345	262.0	3.500	276.6	22.01	6.875	94	-0.67	028.43	0
028	Manganese (mg / kg (ppm))	0045	262.0	20.00	276.6	22.01	6.875	94	-0.66	028.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
028	Manganese (mg / kg (ppm))	0505	263.0	1.800	276.6	22.01	6.875	94	-0.62	028.43	0
028	Manganese (mg / kg (ppm))	0016	265.0	12.00	276.6	22.01	6.875	94	-0.53	028.52	0
028	Manganese (mg / kg (ppm))	0425	265.8	5.100	276.6	22.01	6.875	94	-0.49	028.43	0
028	Manganese (mg / kg (ppm))	0510	267.0	4.000	276.6	22.01	6.875	94	-0.44	028.43	0
028	Manganese (mg / kg (ppm))	0653	269.4	0.0000	276.6	22.01	6.875	94	-0.33	028.41	0
028	Manganese (mg / kg (ppm))	0028	269.4	3.960	276.6	22.01	6.875	94	-0.33	028.43	0
028	Manganese (mg / kg (ppm))	0693	269.7	7.223	276.6	22.01	6.875	94	-0.32	028.42	0
028	Manganese (mg / kg (ppm))	0263	270.0	0.5600	276.6	22.01	6.875	94	-0.30	028.42	0
028	Manganese (mg / kg (ppm))	0035	270.0	2.000	276.6	22.01	6.875	94	-0.30	028.43	0
028	Manganese (mg / kg (ppm))	0353	270.0	8.000	276.6	22.01	6.875	94	-0.30	028.43	0
028	Manganese (mg / kg (ppm))	0098	270.5	3.600	276.6	22.01	6.875	94	-0.28	028.41	0
028	Manganese (mg / kg (ppm))	0033	270.5	1.000	276.6	22.01	6.875	94	-0.28	028.43	0
028	Manganese (mg / kg (ppm))	0148	270.6	12.32	276.6	22.01	6.875	94	-0.28	028.41	0
028	Manganese (mg / kg (ppm))	0619	271.5	1.000	276.6	22.01	6.875	94	-0.23	028.41	0
028	Manganese (mg / kg (ppm))	0868	272.7	0.6800	276.6	22.01	6.875	94	-0.18	028.31	0
028	Manganese (mg / kg (ppm))	0242	273.5	1.000	276.6	22.01	6.875	94	-0.14	028.99	0
028	Manganese (mg / kg (ppm))	2188	274.4	4.720	276.6	22.01	6.875	94	-0.10	028.31	0
028	Manganese (mg / kg (ppm))	0861	276.0	2.000	276.6	22.01	6.875	94	-0.03	028.43	0
028	Manganese (mg / kg (ppm))	0098	276.9	1.500	276.6	22.01	6.875	94	0.01	028.43	0
028	Manganese (mg / kg (ppm))	0009	277.0	14.00	276.6	22.01	6.875	94	0.02	028.42	0
028	Manganese (mg / kg (ppm))	0164	278.0	10.00	276.6	22.01	6.875	94	0.06	028.41	0
028	Manganese (mg / kg (ppm))	0171	278.1	1.900	276.6	22.01	6.875	94	0.06	028.41	0
028	Manganese (mg / kg (ppm))	0187	278.2	0.5400	276.6	22.01	6.875	94	0.07	028.42	0
028	Manganese (mg / kg (ppm))	0074	279.0	4.000	276.6	22.01	6.875	94	0.11	028.41	0
028	Manganese (mg / kg (ppm))	0692	279.0	10.00	276.6	22.01	6.875	94	0.11	028.42	0
028	Manganese (mg / kg (ppm))	0884	279.1	9.400	276.6	22.01	6.875	94	0.11	028.31	0
028	Manganese (mg / kg (ppm))	0685	279.6	2.960	276.6	22.01	6.875	94	0.14	028.42	0
028	Manganese (mg / kg (ppm))	0939	281.3	19.33	276.6	22.01	6.875	94	0.21	028.53	0
028	Manganese (mg / kg (ppm))	0968	281.5	1.000	276.6	22.01	6.875	94	0.22	028.43	0
028	Manganese (mg / kg (ppm))	0035	282.5	1.000	276.6	22.01	6.875	94	0.27	028.42	0
028	Manganese (mg / kg (ppm))	0169	283.0	24.00	276.6	22.01	6.875	94	0.29	028.43	0
028	Manganese (mg / kg (ppm))	0042	283.5	15.00	276.6	22.01	6.875	94	0.31	028.42	0
028	Manganese (mg / kg (ppm))	0354	283.8	1.500	276.6	22.01	6.875	94	0.32	028.31	0
028	Manganese (mg / kg (ppm))	0019	283.8	7.600	276.6	22.01	6.875	94	0.33	028.41	0
028	Manganese (mg / kg (ppm))	0910	285.0	10.00	276.6	22.01	6.875	94	0.38	028.41	0
028	Manganese (mg / kg (ppm))	0948	286.1	7.834	276.6	22.01	6.875	94	0.43	028.33	0
028	Manganese (mg / kg (ppm))	0186	287.5	3.000	276.6	22.01	6.875	94	0.49	028.42	0
028	Manganese (mg / kg (ppm))	0560	287.9	0.4000	276.6	22.01	6.875	94	0.51	028.42	0
028	Manganese (mg / kg (ppm))	0083	288.0	6.000	276.6	22.01	6.875	94	0.52	028.43	0
028	Manganese (mg / kg (ppm))	2196	289.3	0.0000	276.6	22.01	6.875	94	0.58	028.31	0
028	Manganese (mg / kg (ppm))	0682	290.1	0.0000	276.6	22.01	6.875	94	0.61	028.43	0
028	Manganese (mg / kg (ppm))	0047	290.2	5.600	276.6	22.01	6.875	94	0.62	028.52	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
028	Manganese (mg / kg (ppm))	0572	290.5	5.000	276.6	22.01	6.875	94	0.63	028.53	0
028	Manganese (mg / kg (ppm))	0610	291.0	4.000	276.6	22.01	6.875	94	0.65	028.43	0
028	Manganese (mg / kg (ppm))	0723	292.0	0.0000	276.6	22.01	6.875	94	0.70	028.31	0
028	Manganese (mg / kg (ppm))	0229	292.8	2.700	276.6	22.01	6.875	94	0.73	028.41	0
028	Manganese (mg / kg (ppm))	0226	293.3	10.17	276.6	22.01	6.875	94	0.76	028.41	0
028	Manganese (mg / kg (ppm))	0202	293.5	0.1200	276.6	22.01	6.875	94	0.77	028.43	0
028	Manganese (mg / kg (ppm))	0656	293.6	2.780	276.6	22.01	6.875	94	0.77	028.31	0
028	Manganese (mg / kg (ppm))	0849	297.5	5.100	276.6	22.01	6.875	94	0.95	028.31	0
028	Manganese (mg / kg (ppm))	2145	297.8	4.580	276.6	22.01	6.875	94	0.96	028.41	0
028	Manganese (mg / kg (ppm))	0878	298.0	4.000	276.6	22.01	6.875	94	0.97	028.41	0
028	Manganese (mg / kg (ppm))	0357	298.1	4.433	276.6	22.01	6.875	94	0.97	028.42	0
028	Manganese (mg / kg (ppm))	0918	298.3	23.43	276.6	22.01	6.875	94	0.98	028.43	0
028	Manganese (mg / kg (ppm))	0265	300.5	3.000	276.6	22.01	6.875	94	1.08	028.42	0
028	Manganese (mg / kg (ppm))	0726	301.2	0.9000	276.6	22.01	6.875	94	1.11	028.42	0
028	Manganese (mg / kg (ppm))	2144	303.6	7.140	276.6	22.01	6.875	94	1.23	028.31	0
028	Manganese (mg / kg (ppm))	2089	305.1	1.980	276.6	22.01	6.875	94	1.29	028.43	0
028	Manganese (mg / kg (ppm))	0675	305.1	0.7400	276.6	22.01	6.875	94	1.30	028.43	0
028	Manganese (mg / kg (ppm))	0358	309.2	2.430	276.6	22.01	6.875	94	1.48	028.41	0
028	Manganese (mg / kg (ppm))	0590	309.5	7.000	276.6	22.01	6.875	94	1.49	028.99	0
028	Manganese (mg / kg (ppm))	0100	310.0	22.00	276.6	22.01	6.875	94	1.52	028.99	0
028	Manganese (mg / kg (ppm))	2053	311.5	17.00	276.6	22.01	6.875	94	1.58	028.42	0
028	Manganese (mg / kg (ppm))	0186	312.0	8.000	276.6	22.01	6.875	94	1.61	028.52	0
028	Manganese (mg / kg (ppm))	0032	333.5	5.000	276.6	22.01	6.875	94	2.58	028.43	0
028	Manganese (mg / kg (ppm))	0294	336.5	0.7600	276.6	22.01	6.875	94	2.72	028.42	0
028	Manganese (mg / kg (ppm))	2012	621.1	40.99	276.6	22.01	6.875	94	15.65	028.42	0
028	Manganese (mg / kg (ppm))	2128	286.7	92.73	276.6	22.01	6.875	94	0.46	028.32	1
028	Manganese (mg / kg (ppm))	0848	6,259	48.51	276.6	22.01	6.875	94	271.76	028.41	2
031	Phosphorus (%)	0852	1.450	0.1000	2.586	0.1817	0.0512	134	-6.25	031.99	0
031	Phosphorus (%)	0108	1.580	0.0600	2.586	0.1817	0.0512	134	-5.54	031.01	0
031	Phosphorus (%)	0609	1.800	0.0800	2.586	0.1817	0.0512	134	-4.32	031.01	0
031	Phosphorus (%)	0357	2.042	0.0701	2.586	0.1817	0.0512	134	-2.99	031.42	0
031	Phosphorus (%)	0278	2.180	0.0400	2.586	0.1817	0.0512	134	-2.23	031.42	0
031	Phosphorus (%)	0144	2.195	0.0300	2.586	0.1817	0.0512	134	-2.15	031.41	0
031	Phosphorus (%)	0047	2.205	0.0100	2.586	0.1817	0.0512	134	-2.10	031.42	0
031	Phosphorus (%)	0626	2.256	0.0210	2.586	0.1817	0.0512	134	-1.82	031.01	0
031	Phosphorus (%)	0918	2.263	0.0330	2.586	0.1817	0.0512	134	-1.78	031.43	0
031	Phosphorus (%)	0009	2.282	0.1269	2.586	0.1817	0.0512	134	-1.67	031.42	0
031	Phosphorus (%)	0003	2.285	0.0900	2.586	0.1817	0.0512	134	-1.66	031.41	0
031	Phosphorus (%)	0164	2.305	0.0100	2.586	0.1817	0.0512	134	-1.55	031.41	0
031	Phosphorus (%)	0187	2.320	0.0200	2.586	0.1817	0.0512	134	-1.46	031.42	0
031	Phosphorus (%)	0154	2.330	0.0732	2.586	0.1817	0.0512	134	-1.41	031.52	0
031	Phosphorus (%)	0035	2.335	0.0300	2.586	0.1817	0.0512	134	-1.38	031.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	0848	2.345	0.0100	2.586	0.1817	0.0512	134	-1.32	031.41	0
031	Phosphorus (%)	0536	2.355	0.0500	2.586	0.1817	0.0512	134	-1.27	031.06	0
031	Phosphorus (%)	0405	2.360	0.0000	2.586	0.1817	0.0512	134	-1.24	031.41	0
031	Phosphorus (%)	0619	2.380	0.1800	2.586	0.1817	0.0512	134	-1.13	031.41	0
031	Phosphorus (%)	0345	2.385	0.0900	2.586	0.1817	0.0512	134	-1.10	031.43	0
031	Phosphorus (%)	0874	2.387	0.0270	2.586	0.1817	0.0512	134	-1.10	031.01	0
031	Phosphorus (%)	0019	2.390	0.0400	2.586	0.1817	0.0512	134	-1.08	031.41	0
031	Phosphorus (%)	0520	2.390	0.0800	2.586	0.1817	0.0512	134	-1.08	031.41	0
031	Phosphorus (%)	2199	2.395	0.0100	2.586	0.1817	0.0512	134	-1.05	031.01	0
031	Phosphorus (%)	0027	2.401	0.1110	2.586	0.1817	0.0512	134	-1.02	031.43	0
031	Phosphorus (%)	0870	2.402	0.1050	2.586	0.1817	0.0512	134	-1.01	031.42	0
031	Phosphorus (%)	0720	2.410	0.1000	2.586	0.1817	0.0512	134	-0.97	031.41	0
031	Phosphorus (%)	0674	2.430	0.0000	2.586	0.1817	0.0512	134	-0.86	031.01	0
031	Phosphorus (%)	0089	2.430	0.0000	2.586	0.1817	0.0512	134	-0.86	031.43	0
031	Phosphorus (%)	2012	2.440	0.1800	2.586	0.1817	0.0512	134	-0.80	031.42	0
031	Phosphorus (%)	0010	2.450	0.0200	2.586	0.1817	0.0512	134	-0.75	031.43	0
031	Phosphorus (%)	0687	2.455	0.2900	2.586	0.1817	0.0512	134	-0.72	031.01	0
031	Phosphorus (%)	0123	2.455	0.0300	2.586	0.1817	0.0512	134	-0.72	031.41	0
031	Phosphorus (%)	0001	2.457	0.0590	2.586	0.1817	0.0512	134	-0.71	031.03	0
031	Phosphorus (%)	2055	2.460	0.2226	2.586	0.1817	0.0512	134	-0.69	031.53	0
031	Phosphorus (%)	0035	2.480	0.0800	2.586	0.1817	0.0512	134	-0.58	031.42	0
031	Phosphorus (%)	0205	2.480	0.0400	2.586	0.1817	0.0512	134	-0.58	031.42	0
031	Phosphorus (%)	2200	2.500	0.0000	2.586	0.1817	0.0512	134	-0.47	031.01	0
031	Phosphorus (%)	0016	2.510	0.2400	2.586	0.1817	0.0512	134	-0.42	031.01	0
031	Phosphorus (%)	0553	2.510	0.0000	2.586	0.1817	0.0512	134	-0.42	031.41	0
031	Phosphorus (%)	0294	2.510	0.0400	2.586	0.1817	0.0512	134	-0.42	031.42	0
031	Phosphorus (%)	0202	2.512	0.0250	2.586	0.1817	0.0512	134	-0.41	031.43	0
031	Phosphorus (%)	0964	2.512	0.1618	2.586	0.1817	0.0512	134	-0.41	031.43	0
031	Phosphorus (%)	0693	2.525	0.0010	2.586	0.1817	0.0512	134	-0.34	031.42	0
031	Phosphorus (%)	0881	2.525	0.0060	2.586	0.1817	0.0512	134	-0.33	031.01	0
031	Phosphorus (%)	0610	2.525	0.0300	2.586	0.1817	0.0512	134	-0.33	031.43	0
031	Phosphorus (%)	0171	2.530	0.0000	2.586	0.1817	0.0512	134	-0.31	031.41	0
031	Phosphorus (%)	0045	2.530	0.0200	2.586	0.1817	0.0512	134	-0.31	031.42	0
031	Phosphorus (%)	0142	2.538	0.0777	2.586	0.1817	0.0512	134	-0.26	031.01	0
031	Phosphorus (%)	0726	2.539	0.0100	2.586	0.1817	0.0512	134	-0.26	031.42	0
031	Phosphorus (%)	2215	2.540	0.0000	2.586	0.1817	0.0512	134	-0.25	031.01	0
031	Phosphorus (%)	0676	2.540	0.0200	2.586	0.1817	0.0512	134	-0.25	031.99	0
031	Phosphorus (%)	0100	2.545	0.0700	2.586	0.1817	0.0512	134	-0.22	031.99	0
031	Phosphorus (%)	0897	2.550	0.0000	2.586	0.1817	0.0512	134	-0.20	031.01	0
031	Phosphorus (%)	0042	2.550	0.2800	2.586	0.1817	0.0512	134	-0.20	031.42	0
031	Phosphorus (%)	0098	2.553	0.1690	2.586	0.1817	0.0512	134	-0.18	031.41	0
031	Phosphorus (%)	0186	2.565	0.1500	2.586	0.1817	0.0512	134	-0.11	031.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	0651	2.573	0.0100	2.586	0.1817	0.0512	134	-0.07	031.01	0
031	Phosphorus (%)	2062	2.580	0.0200	2.586	0.1817	0.0512	134	-0.03	031.01	0
031	Phosphorus (%)	0685	2.585	0.0100	2.586	0.1817	0.0512	134	0.00	031.41	0
031	Phosphorus (%)	2181	2.590	0.0600	2.586	0.1817	0.0512	134	0.02	031.01	0
031	Phosphorus (%)	0425	2.595	0.0100	2.586	0.1817	0.0512	134	0.05	031.43	0
031	Phosphorus (%)	2009	2.596	0.0031	2.586	0.1817	0.0512	134	0.05	031.01	0
031	Phosphorus (%)	2193	2.600	0.0200	2.586	0.1817	0.0512	134	0.08	031.01	0
031	Phosphorus (%)	0186	2.600	0.0600	2.586	0.1817	0.0512	134	0.08	031.52	0
031	Phosphorus (%)	0242	2.600	0.0200	2.586	0.1817	0.0512	134	0.08	031.99	0
031	Phosphorus (%)	0148	2.605	0.0100	2.586	0.1817	0.0512	134	0.11	031.41	0
031	Phosphorus (%)	0037	2.610	0.0400	2.586	0.1817	0.0512	134	0.13	031.42	0
031	Phosphorus (%)	2188	2.610	0.0200	2.586	0.1817	0.0512	134	0.13	031.01	0
031	Phosphorus (%)	2128	2.615	0.1280	2.586	0.1817	0.0512	134	0.16	031.01	0
031	Phosphorus (%)	2089	2.615	0.0500	2.586	0.1817	0.0512	134	0.16	031.43	0
031	Phosphorus (%)	2192	2.620	0.0000	2.586	0.1817	0.0512	134	0.19	031.01	0
031	Phosphorus (%)	0033	2.620	0.0800	2.586	0.1817	0.0512	134	0.19	031.43	0
031	Phosphorus (%)	0849	2.621	0.0020	2.586	0.1817	0.0512	134	0.19	031.01	0
031	Phosphorus (%)	0190	2.630	0.0800	2.586	0.1817	0.0512	134	0.24	031.42	0
031	Phosphorus (%)	0049	2.640	0.0600	2.586	0.1817	0.0512	134	0.30	031.41	0
031	Phosphorus (%)	0175	2.640	0.0800	2.586	0.1817	0.0512	134	0.30	031.01	0
031	Phosphorus (%)	0656	2.640	0.0000	2.586	0.1817	0.0512	134	0.30	031.01	0
031	Phosphorus (%)	0868	2.645	0.1700	2.586	0.1817	0.0512	134	0.33	031.01	0
031	Phosphorus (%)	2006	2.650	0.0190	2.586	0.1817	0.0512	134	0.35	031.01	0
031	Phosphorus (%)	0723	2.650	0.0000	2.586	0.1817	0.0512	134	0.35	031.01	0
031	Phosphorus (%)	2196	2.650	0.0000	2.586	0.1817	0.0512	134	0.35	031.01	0
031	Phosphorus (%)	0169	2.650	0.0200	2.586	0.1817	0.0512	134	0.35	031.03	0
031	Phosphorus (%)	0036	2.650	0.0011	2.586	0.1817	0.0512	134	0.36	031.03	0
031	Phosphorus (%)	0948	2.654	0.0050	2.586	0.1817	0.0512	134	0.37	031.01	0
031	Phosphorus (%)	0560	2.658	0.0671	2.586	0.1817	0.0512	134	0.40	031.42	0
031	Phosphorus (%)	0629	2.660	0.1200	2.586	0.1817	0.0512	134	0.41	031.01	0
031	Phosphorus (%)	0689	2.660	0.0400	2.586	0.1817	0.0512	134	0.41	031.01	0
031	Phosphorus (%)	0407	2.663	0.0434	2.586	0.1817	0.0512	134	0.43	031.41	0
031	Phosphorus (%)	2190	2.674	0.0005	2.586	0.1817	0.0512	134	0.49	031.01	0
031	Phosphorus (%)	2145	2.674	0.0132	2.586	0.1817	0.0512	134	0.49	031.41	0
031	Phosphorus (%)	0861	2.675	0.0500	2.586	0.1817	0.0512	134	0.49	031.43	0
031	Phosphorus (%)	0263	2.679	0.0040	2.586	0.1817	0.0512	134	0.51	031.42	0
031	Phosphorus (%)	0354	2.680	0.0000	2.586	0.1817	0.0512	134	0.52	031.01	0
031	Phosphorus (%)	0511	2.680	0.0000	2.586	0.1817	0.0512	134	0.52	031.01	0
031	Phosphorus (%)	0683	2.680	0.0800	2.586	0.1817	0.0512	134	0.52	031.01	0
031	Phosphorus (%)	2053	2.680	0.0000	2.586	0.1817	0.0512	134	0.52	031.02	0
031	Phosphorus (%)	0229	2.690	0.0400	2.586	0.1817	0.0512	134	0.57	031.41	0
031	Phosphorus (%)	0682	2.690	0.0000	2.586	0.1817	0.0512	134	0.57	031.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	0675	2.690	0.0200	2.586	0.1817	0.0512	134	0.57	031.43	0
031	Phosphorus (%)	2198	2.695	0.0300	2.586	0.1817	0.0512	134	0.60	031.01	0
031	Phosphorus (%)	0968	2.697	0.0040	2.586	0.1817	0.0512	134	0.61	031.43	0
031	Phosphorus (%)	0590	2.700	0.0000	2.586	0.1817	0.0512	134	0.63	031.01	0
031	Phosphorus (%)	2195	2.710	0.0000	2.586	0.1817	0.0512	134	0.68	031.01	0
031	Phosphorus (%)	0510	2.710	0.0000	2.586	0.1817	0.0512	134	0.68	031.43	0
031	Phosphorus (%)	2146	2.711	0.0353	2.586	0.1817	0.0512	134	0.69	031.51	0
031	Phosphorus (%)	0036	2.713	0.0347	2.586	0.1817	0.0512	134	0.70	031.43	0
031	Phosphorus (%)	0505	2.725	0.0100	2.586	0.1817	0.0512	134	0.77	031.02	0
031	Phosphorus (%)	0893	2.730	0.0000	2.586	0.1817	0.0512	134	0.79	031.00	0
031	Phosphorus (%)	0878	2.730	0.0200	2.586	0.1817	0.0512	134	0.79	031.01	0
031	Phosphorus (%)	2022	2.735	0.0100	2.586	0.1817	0.0512	134	0.82	031.01	0
031	Phosphorus (%)	0553	2.735	0.0100	2.586	0.1817	0.0512	134	0.82	031.53	0
031	Phosphorus (%)	0098	2.736	0.0710	2.586	0.1817	0.0512	134	0.82	031.43	0
031	Phosphorus (%)	0910	2.740	0.2600	2.586	0.1817	0.0512	134	0.85	031.41	0
031	Phosphorus (%)	0653	2.742	0.0040	2.586	0.1817	0.0512	134	0.86	031.41	0
031	Phosphorus (%)	2126	2.755	0.0100	2.586	0.1817	0.0512	134	0.93	031.01	0
031	Phosphorus (%)	0572	2.755	0.0300	2.586	0.1817	0.0512	134	0.93	031.53	0
031	Phosphorus (%)	0885	2.774	0.0117	2.586	0.1817	0.0512	134	1.04	031.01	0
031	Phosphorus (%)	0939	2.790	0.0600	2.586	0.1817	0.0512	134	1.12	031.53	0
031	Phosphorus (%)	0504	2.805	0.0300	2.586	0.1817	0.0512	134	1.21	031.03	0
031	Phosphorus (%)	2109	2.807	0.0190	2.586	0.1817	0.0512	134	1.22	031.01	0
031	Phosphorus (%)	0265	2.820	0.0400	2.586	0.1817	0.0512	134	1.29	031.42	0
031	Phosphorus (%)	0032	2.820	0.0200	2.586	0.1817	0.0512	134	1.29	031.43	0
031	Phosphorus (%)	0083	2.835	0.1100	2.586	0.1817	0.0512	134	1.37	031.43	0
031	Phosphorus (%)	0028	2.840	0.1400	2.586	0.1817	0.0512	134	1.40	031.43	0
031	Phosphorus (%)	0598	2.880	0.0110	2.586	0.1817	0.0512	134	1.62	031.41	0
031	Phosphorus (%)	0074	2.880	0.0000	2.586	0.1817	0.0512	134	1.62	031.41	0
031	Phosphorus (%)	0353	2.895	0.0500	2.586	0.1817	0.0512	134	1.70	031.43	0
031	Phosphorus (%)	0226	2.900	0.0800	2.586	0.1817	0.0512	134	1.73	031.41	0
031	Phosphorus (%)	2129	2.910	0.0200	2.586	0.1817	0.0512	134	1.78	031.42	0
031	Phosphorus (%)	0018	2.955	0.3100	2.586	0.1817	0.0512	134	2.03	031.01	0
031	Phosphorus (%)	2031	2.970	0.0610	2.586	0.1817	0.0512	134	2.11	031.01	0
031	Phosphorus (%)	0358	3.120	0.0800	2.586	0.1817	0.0512	134	2.94	031.41	0
031	Phosphorus (%)	2223	3.255	0.0100	2.586	0.1817	0.0512	134	3.68	031.01	0
031	Phosphorus (%)	0692	2.520	0.3800	2.586	0.1817	0.0512	134	-0.36	031.42	1
031	Phosphorus (%)	0884	2.600	0.5400	2.586	0.1817	0.0512	134	0.08	031.01	1
031	Phosphorus (%)	0300	3.055	0.3500	2.586	0.1817	0.0512	134	2.58	031.99	1
032	Potassium (%)	2146	0.4453	0.0027	1.443	0.0803	0.0418	101	-12.42	032.31	0
032	Potassium (%)	2109	1.216	0.0110	1.443	0.0803	0.0418	101	-2.83	032.08	0
032	Potassium (%)	0609	1.220	0.0600	1.443	0.0803	0.0418	101	-2.77	032.32	0
032	Potassium (%)	0405	1.240	0.0000	1.443	0.0803	0.0418	101	-2.53	032.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
032	Potassium (%)	0169	1.270	0.0600	1.443	0.0803	0.0418	101	-2.15	032.32	0
032	Potassium (%)	0918	1.283	0.1020	1.443	0.0803	0.0418	101	-1.99	032.43	0
032	Potassium (%)	2089	1.295	0.1100	1.443	0.0803	0.0418	101	-1.84	032.43	0
032	Potassium (%)	0016	1.300	0.0400	1.443	0.0803	0.0418	101	-1.78	032.52	0
032	Potassium (%)	0098	1.315	0.0100	1.443	0.0803	0.0418	101	-1.59	032.41	0
032	Potassium (%)	2129	1.320	0.0200	1.443	0.0803	0.0418	101	-1.53	032.42	0
032	Potassium (%)	0144	1.330	0.0200	1.443	0.0803	0.0418	101	-1.40	032.41	0
032	Potassium (%)	0619	1.340	0.0400	1.443	0.0803	0.0418	101	-1.28	032.41	0
032	Potassium (%)	0881	1.355	0.0590	1.443	0.0803	0.0418	101	-1.10	032.99	0
032	Potassium (%)	0874	1.355	0.0100	1.443	0.0803	0.0418	101	-1.09	032.31	0
032	Potassium (%)	0553	1.355	0.0100	1.443	0.0803	0.0418	101	-1.09	032.41	0
032	Potassium (%)	0278	1.360	0.0400	1.443	0.0803	0.0418	101	-1.03	032.42	0
032	Potassium (%)	0035	1.360	0.0200	1.443	0.0803	0.0418	101	-1.03	032.32	0
032	Potassium (%)	0035	1.360	0.0200	1.443	0.0803	0.0418	101	-1.03	032.42	0
032	Potassium (%)	0536	1.365	0.0100	1.443	0.0803	0.0418	101	-0.97	032.31	0
032	Potassium (%)	0510	1.370	0.0000	1.443	0.0803	0.0418	101	-0.91	032.43	0
032	Potassium (%)	0098	1.380	0.0400	1.443	0.0803	0.0418	101	-0.78	032.43	0
032	Potassium (%)	0505	1.385	0.0300	1.443	0.0803	0.0418	101	-0.72	032.43	0
032	Potassium (%)	0939	1.385	0.0900	1.443	0.0803	0.0418	101	-0.72	032.53	0
032	Potassium (%)	0653	1.388	0.0040	1.443	0.0803	0.0418	101	-0.68	032.41	0
032	Potassium (%)	0964	1.393	0.0768	1.443	0.0803	0.0418	101	-0.62	032.43	0
032	Potassium (%)	0629	1.395	0.0500	1.443	0.0803	0.0418	101	-0.60	032.31	0
032	Potassium (%)	0035	1.395	0.0300	1.443	0.0803	0.0418	101	-0.60	032.43	0
032	Potassium (%)	0175	1.400	0.0200	1.443	0.0803	0.0418	101	-0.53	032.31	0
032	Potassium (%)	0045	1.400	0.0200	1.443	0.0803	0.0418	101	-0.53	032.42	0
032	Potassium (%)	0042	1.405	0.0900	1.443	0.0803	0.0418	101	-0.47	032.42	0
032	Potassium (%)	0693	1.412	0.0320	1.443	0.0803	0.0418	101	-0.38	032.42	0
032	Potassium (%)	2055	1.419	0.0703	1.443	0.0803	0.0418	101	-0.30	032.53	0
032	Potassium (%)	0870	1.420	0.0540	1.443	0.0803	0.0418	101	-0.28	032.42	0
032	Potassium (%)	0425	1.420	0.0200	1.443	0.0803	0.0418	101	-0.28	032.43	0
032	Potassium (%)	0682	1.420	0.0000	1.443	0.0803	0.0418	101	-0.28	032.43	0
032	Potassium (%)	0720	1.425	0.0100	1.443	0.0803	0.0418	101	-0.22	032.41	0
032	Potassium (%)	0590	1.425	0.0300	1.443	0.0803	0.0418	101	-0.22	032.99	0
032	Potassium (%)	0910	1.425	0.1500	1.443	0.0803	0.0418	101	-0.22	032.41	0
032	Potassium (%)	0001	1.427	0.1200	1.443	0.0803	0.0418	101	-0.20	032.99	0
032	Potassium (%)	0407	1.428	0.0634	1.443	0.0803	0.0418	101	-0.19	032.41	0
032	Potassium (%)	2053	1.430	0.0200	1.443	0.0803	0.0418	101	-0.16	032.31	0
032	Potassium (%)	0358	1.430	0.0200	1.443	0.0803	0.0418	101	-0.16	032.41	0
032	Potassium (%)	0033	1.430	0.0000	1.443	0.0803	0.0418	101	-0.16	032.43	0
032	Potassium (%)	2022	1.435	0.0100	1.443	0.0803	0.0418	101	-0.10	032.31	0
032	Potassium (%)	0353	1.435	0.0300	1.443	0.0803	0.0418	101	-0.10	032.43	0
032	Potassium (%)	0504	1.440	0.0000	1.443	0.0803	0.0418	101	-0.03	032.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
032	Potassium (%)	0345	1.440	0.0200	1.443	0.0803	0.0418	101	-0.03	032.43	0
032	Potassium (%)	0006	1.444	0.0050	1.443	0.0803	0.0418	101	0.01	032.43	0
032	Potassium (%)	0357	1.444	0.0185	1.443	0.0803	0.0418	101	0.02	032.42	0
032	Potassium (%)	0868	1.445	0.0100	1.443	0.0803	0.0418	101	0.03	032.31	0
032	Potassium (%)	0019	1.445	0.0500	1.443	0.0803	0.0418	101	0.03	032.41	0
032	Potassium (%)	0511	1.445	0.0100	1.443	0.0803	0.0418	101	0.03	032.41	0
032	Potassium (%)	0047	1.445	0.0500	1.443	0.0803	0.0418	101	0.03	032.42	0
032	Potassium (%)	0187	1.445	0.0100	1.443	0.0803	0.0418	101	0.03	032.42	0
032	Potassium (%)	0100	1.445	0.0500	1.443	0.0803	0.0418	101	0.03	032.99	0
032	Potassium (%)	0148	1.450	0.0400	1.443	0.0803	0.0418	101	0.09	032.41	0
032	Potassium (%)	0229	1.450	0.0000	1.443	0.0803	0.0418	101	0.09	032.41	0
032	Potassium (%)	0520	1.450	0.0000	1.443	0.0803	0.0418	101	0.09	032.41	0
032	Potassium (%)	0242	1.450	0.0000	1.443	0.0803	0.0418	101	0.09	032.99	0
032	Potassium (%)	0354	1.455	0.0100	1.443	0.0803	0.0418	101	0.15	032.31	0
032	Potassium (%)	0948	1.458	0.0110	1.443	0.0803	0.0418	101	0.18	032.31	0
032	Potassium (%)	0074	1.460	0.0200	1.443	0.0803	0.0418	101	0.21	032.41	0
032	Potassium (%)	0164	1.460	0.1000	1.443	0.0803	0.0418	101	0.21	032.41	0
032	Potassium (%)	0683	1.465	0.0100	1.443	0.0803	0.0418	101	0.28	032.31	0
032	Potassium (%)	0010	1.465	0.0500	1.443	0.0803	0.0418	101	0.28	032.43	0
032	Potassium (%)	0878	1.470	0.0200	1.443	0.0803	0.0418	101	0.34	032.41	0
032	Potassium (%)	0861	1.470	0.0400	1.443	0.0803	0.0418	101	0.34	032.43	0
032	Potassium (%)	0009	1.473	0.0375	1.443	0.0803	0.0418	101	0.37	032.42	0
032	Potassium (%)	0202	1.474	0.0110	1.443	0.0803	0.0418	101	0.38	032.43	0
032	Potassium (%)	2195	1.475	0.0100	1.443	0.0803	0.0418	101	0.40	032.31	0
032	Potassium (%)	0049	1.475	0.1100	1.443	0.0803	0.0418	101	0.40	032.41	0
032	Potassium (%)	0205	1.475	0.0300	1.443	0.0803	0.0418	101	0.40	032.42	0
032	Potassium (%)	0656	1.485	0.0100	1.443	0.0803	0.0418	101	0.53	032.31	0
032	Potassium (%)	0560	1.491	0.0473	1.443	0.0803	0.0418	101	0.60	032.42	0
032	Potassium (%)	0598	1.492	0.0320	1.443	0.0803	0.0418	101	0.61	032.41	0
032	Potassium (%)	0083	1.495	0.0100	1.443	0.0803	0.0418	101	0.65	032.43	0
032	Potassium (%)	0726	1.497	0.0020	1.443	0.0803	0.0418	101	0.67	032.42	0
032	Potassium (%)	0265	1.500	0.0600	1.443	0.0803	0.0418	101	0.71	032.42	0
032	Potassium (%)	0553	1.500	0.0400	1.443	0.0803	0.0418	101	0.71	032.53	0
032	Potassium (%)	0612	1.505	0.0100	1.443	0.0803	0.0418	101	0.77	032.33	0
032	Potassium (%)	0692	1.505	0.1500	1.443	0.0803	0.0418	101	0.77	032.42	0
032	Potassium (%)	0186	1.505	0.0300	1.443	0.0803	0.0418	101	0.77	032.52	0
032	Potassium (%)	0142	1.507	0.0388	1.443	0.0803	0.0418	101	0.80	032.31	0
032	Potassium (%)	0263	1.513	0.0364	1.443	0.0803	0.0418	101	0.87	032.42	0
032	Potassium (%)	0171	1.513	0.0000	1.443	0.0803	0.0418	101	0.87	032.41	0
032	Potassium (%)	0226	1.515	0.1300	1.443	0.0803	0.0418	101	0.90	032.41	0
032	Potassium (%)	0003	1.520	0.1600	1.443	0.0803	0.0418	101	0.96	032.41	0
032	Potassium (%)	0036	1.528	0.0138	1.443	0.0803	0.0418	101	1.05	032.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
032	Potassium (%)	0610	1.530	0.0000	1.443	0.0803	0.0418	101	1.09	032.43	0
032	Potassium (%)	0190	1.535	0.0300	1.443	0.0803	0.0418	101	1.15	032.42	0
032	Potassium (%)	2012	1.565	0.0900	1.443	0.0803	0.0418	101	1.52	032.42	0
032	Potassium (%)	0968	1.572	0.0030	1.443	0.0803	0.0418	101	1.60	032.43	0
032	Potassium (%)	0037	1.575	0.0300	1.443	0.0803	0.0418	101	1.65	032.42	0
032	Potassium (%)	0186	1.580	0.1200	1.443	0.0803	0.0418	101	1.71	032.42	0
032	Potassium (%)	0675	1.585	0.0500	1.443	0.0803	0.0418	101	1.77	032.43	0
032	Potassium (%)	0154	1.587	0.1558	1.443	0.0803	0.0418	101	1.80	032.52	0
032	Potassium (%)	0572	1.590	0.2400	1.443	0.0803	0.0418	101	1.83	032.53	0
032	Potassium (%)	0089	1.590	0.0000	1.443	0.0803	0.0418	101	1.83	032.43	0
032	Potassium (%)	0032	1.610	0.0000	1.443	0.0803	0.0418	101	2.08	032.43	0
032	Potassium (%)	0028	1.680	0.2400	1.443	0.0803	0.0418	101	2.95	032.43	0
032	Potassium (%)	0294	2.750	0.0400	1.443	0.0803	0.0418	101	16.28	032.42	0
032	Potassium (%)	0848	14,883	6.940	1.443	0.0803	0.0418	101	185332.41	032.41	2
033	Salt as chloride (%)	0885	0.4523	0.0069	1.919	0.1214	0.0294	69	-12.08	033.00	0
033	Salt as chloride (%)	2193	0.5950	0.0700	1.919	0.1214	0.0294	69	-10.91	033.00	0
033	Salt as chloride (%)	0881	0.6015	0.0130	1.919	0.1214	0.0294	69	-10.85	033.00	0
033	Salt as chloride (%)	0358	1.035	0.0500	1.919	0.1214	0.0294	69	-7.28	033.99	0
033	Salt as chloride (%)	0027	1.065	0.0130	1.919	0.1214	0.0294	69	-7.04	033.99	0
033	Salt as chloride (%)	0536	1.365	0.0700	1.919	0.1214	0.0294	69	-4.56	033.99	0
033	Salt as chloride (%)	0675	1.410	0.1400	1.919	0.1214	0.0294	69	-4.19	033.00	0
033	Salt as chloride (%)	0171	1.450	0.0000	1.919	0.1214	0.0294	69	-3.86	033.00	0
033	Salt as chloride (%)	0539	1.500	0.0400	1.919	0.1214	0.0294	69	-3.45	033.00	0
033	Salt as chloride (%)	0897	1.520	0.0000	1.919	0.1214	0.0294	69	-3.29	033.00	0
033	Salt as chloride (%)	0656	1.600	0.0000	1.919	0.1214	0.0294	69	-2.63	033.03	0
033	Salt as chloride (%)	0868	1.720	0.0200	1.919	0.1214	0.0294	69	-1.64	033.05	0
033	Salt as chloride (%)	0265	1.745	0.0300	1.919	0.1214	0.0294	69	-1.43	033.03	0
033	Salt as chloride (%)	0893	1.810	0.0000	1.919	0.1214	0.0294	69	-0.90	033.00	0
033	Salt as chloride (%)	0726	1.820	0.0000	1.919	0.1214	0.0294	69	-0.82	033.03	0
033	Salt as chloride (%)	0505	1.835	0.0700	1.919	0.1214	0.0294	69	-0.69	033.03	0
033	Salt as chloride (%)	2129	1.835	0.0300	1.919	0.1214	0.0294	69	-0.69	033.05	0
033	Salt as chloride (%)	0190	1.840	0.0000	1.919	0.1214	0.0294	69	-0.65	033.03	0
033	Salt as chloride (%)	0045	1.865	0.0300	1.919	0.1214	0.0294	69	-0.44	033.00	0
033	Salt as chloride (%)	0878	1.865	0.0500	1.919	0.1214	0.0294	69	-0.44	033.00	0
033	Salt as chloride (%)	2076	1.870	0.0000	1.919	0.1214	0.0294	69	-0.40	033.00	0
033	Salt as chloride (%)	0353	1.875	0.0300	1.919	0.1214	0.0294	69	-0.36	033.00	0
033	Salt as chloride (%)	0229	1.890	0.0200	1.919	0.1214	0.0294	69	-0.24	033.01	0
033	Salt as chloride (%)	2009	1.895	0.0028	1.919	0.1214	0.0294	69	-0.20	033.00	0
033	Salt as chloride (%)	0309	1.898	0.0280	1.919	0.1214	0.0294	69	-0.17	033.00	0
033	Salt as chloride (%)	2022	1.900	0.0000	1.919	0.1214	0.0294	69	-0.16	033.00	0
033	Salt as chloride (%)	0853	1.900	0.2000	1.919	0.1214	0.0294	69	-0.16	033.99	0
033	Salt as chloride (%)	0910	1.900	0.0200	1.919	0.1214	0.0294	69	-0.16	033.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
033	Salt as chloride (%)	0510	1.910	0.1000	1.919	0.1214	0.0294	69	-0.07	033.01	0
033	Salt as chloride (%)	2196	1.920	0.0000	1.919	0.1214	0.0294	69	0.01	033.00	0
033	Salt as chloride (%)	0425	1.945	0.0700	1.919	0.1214	0.0294	69	0.21	033.01	0
033	Salt as chloride (%)	0169	1.950	0.0000	1.919	0.1214	0.0294	69	0.26	033.00	0
033	Salt as chloride (%)	2145	1.950	0.0200	1.919	0.1214	0.0294	69	0.26	033.99	0
033	Salt as chloride (%)	0083	1.960	0.0000	1.919	0.1214	0.0294	69	0.34	033.01	0
033	Salt as chloride (%)	0504	1.965	0.1100	1.919	0.1214	0.0294	69	0.38	033.00	0
033	Salt as chloride (%)	0004	1.965	0.0300	1.919	0.1214	0.0294	69	0.38	033.01	0
033	Salt as chloride (%)	2062	1.970	0.0200	1.919	0.1214	0.0294	69	0.42	033.00	0
033	Salt as chloride (%)	0682	1.970	0.0000	1.919	0.1214	0.0294	69	0.42	033.99	0
033	Salt as chloride (%)	0948	1.975	0.0300	1.919	0.1214	0.0294	69	0.46	033.01	0
033	Salt as chloride (%)	0100	1.980	0.0000	1.919	0.1214	0.0294	69	0.50	033.01	0
033	Salt as chloride (%)	0693	1.985	0.0430	1.919	0.1214	0.0294	69	0.54	033.00	0
033	Salt as chloride (%)	2144	1.985	0.0300	1.919	0.1214	0.0294	69	0.54	033.01	0
033	Salt as chloride (%)	0205	1.990	0.1000	1.919	0.1214	0.0294	69	0.59	033.01	0
033	Salt as chloride (%)	0226	1.990	0.0200	1.919	0.1214	0.0294	69	0.59	033.01	0
033	Salt as chloride (%)	2109	1.992	0.0200	1.919	0.1214	0.0294	69	0.60	033.01	0
033	Salt as chloride (%)	0939	1.995	0.0300	1.919	0.1214	0.0294	69	0.63	033.01	0
033	Salt as chloride (%)	0175	1.995	0.0100	1.919	0.1214	0.0294	69	0.63	033.01	0
033	Salt as chloride (%)	0683	1.995	0.0100	1.919	0.1214	0.0294	69	0.63	033.01	0
033	Salt as chloride (%)	0689	1.995	0.0100	1.919	0.1214	0.0294	69	0.63	033.05	0
033	Salt as chloride (%)	0183	2.000	0.0000	1.919	0.1214	0.0294	69	0.67	033.00	0
033	Salt as chloride (%)	0354	2.000	0.0000	1.919	0.1214	0.0294	69	0.67	033.01	0
033	Salt as chloride (%)	0861	2.000	0.0000	1.919	0.1214	0.0294	69	0.67	033.01	0
033	Salt as chloride (%)	0964	2.000	0.0580	1.919	0.1214	0.0294	69	0.67	033.01	0
033	Salt as chloride (%)	0144	2.000	0.0000	1.919	0.1214	0.0294	69	0.67	033.03	0
033	Salt as chloride (%)	0098	2.005	0.0100	1.919	0.1214	0.0294	69	0.71	033.01	0
033	Salt as chloride (%)	2190	2.007	0.0152	1.919	0.1214	0.0294	69	0.73	033.00	0
033	Salt as chloride (%)	0644	2.008	0.0233	1.919	0.1214	0.0294	69	0.73	033.01	0
033	Salt as chloride (%)	0164	2.010	0.0000	1.919	0.1214	0.0294	69	0.75	033.01	0
033	Salt as chloride (%)	0242	2.010	0.0600	1.919	0.1214	0.0294	69	0.75	033.01	0
033	Salt as chloride (%)	0874	2.010	0.0000	1.919	0.1214	0.0294	69	0.75	033.01	0
033	Salt as chloride (%)	0001	2.015	0.0100	1.919	0.1214	0.0294	69	0.79	033.01	0
033	Salt as chloride (%)	0148	2.017	0.0020	1.919	0.1214	0.0294	69	0.81	033.01	0
033	Salt as chloride (%)	0278	2.020	0.0200	1.919	0.1214	0.0294	69	0.83	033.01	0
033	Salt as chloride (%)	0629	2.020	0.0000	1.919	0.1214	0.0294	69	0.83	033.01	0
033	Salt as chloride (%)	0407	2.028	0.0030	1.919	0.1214	0.0294	69	0.89	033.01	0
033	Salt as chloride (%)	0019	2.045	0.0100	1.919	0.1214	0.0294	69	1.04	033.01	0
033	Salt as chloride (%)	2073	2.065	0.0300	1.919	0.1214	0.0294	69	1.20	033.99	0
033	Salt as chloride (%)	0596	2.080	0.0000	1.919	0.1214	0.0294	69	1.33	033.01	0
033	Salt as chloride (%)	0003	2.200	0.2000	1.919	0.1214	0.0294	69	2.31	033.01	0
033	Salt as chloride (%)	2181	1.800	0.3000	1.919	0.1214	0.0294	69	-0.98	033.99	1

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
033	Salt as chloride (%)	0511	1.870	0.2400	1.919	0.1214	0.0294	69	-0.40	033.00	1
034	Selenium (mg / kg (ppm))	0848	2.400	0.0200	14.40	1.907	0.4013	27	-6.29	034.41	0
034	Selenium (mg / kg (ppm))	0032	8.260	0.0600	14.40	1.907	0.4013	27	-3.22	034.53	0
034	Selenium (mg / kg (ppm))	0045	10.34	0.7300	14.40	1.907	0.4013	27	-2.13	034.04	0
034	Selenium (mg / kg (ppm))	2055	11.71	0.1800	14.40	1.907	0.4013	27	-1.41	034.53	0
034	Selenium (mg / kg (ppm))	0955	11.85	1.700	14.40	1.907	0.4013	27	-1.34	034.43	0
034	Selenium (mg / kg (ppm))	0148	13.19	0.1800	14.40	1.907	0.4013	27	-0.64	034.41	0
034	Selenium (mg / kg (ppm))	0190	13.34	0.1100	14.40	1.907	0.4013	27	-0.56	034.42	0
034	Selenium (mg / kg (ppm))	0619	13.40	0.6000	14.40	1.907	0.4013	27	-0.53	034.41	0
034	Selenium (mg / kg (ppm))	0027	13.91	0.7390	14.40	1.907	0.4013	27	-0.26	034.43	0
034	Selenium (mg / kg (ppm))	0186	14.00	0.0000	14.40	1.907	0.4013	27	-0.21	034.42	0
034	Selenium (mg / kg (ppm))	0964	14.12	0.3491	14.40	1.907	0.4013	27	-0.15	034.43	0
034	Selenium (mg / kg (ppm))	0016	14.30	0.2000	14.40	1.907	0.4013	27	-0.05	034.52	0
034	Selenium (mg / kg (ppm))	0164	14.40	0.4000	14.40	1.907	0.4013	27	0.00	034.53	0
034	Selenium (mg / kg (ppm))	0033	14.50	0.4000	14.40	1.907	0.4013	27	0.05	034.53	0
034	Selenium (mg / kg (ppm))	0968	14.97	0.0770	14.40	1.907	0.4013	27	0.30	034.43	0
034	Selenium (mg / kg (ppm))	0553	15.20	0.0000	14.40	1.907	0.4013	27	0.42	034.53	0
034	Selenium (mg / kg (ppm))	0098	15.30	0.2000	14.40	1.907	0.4013	27	0.47	034.53	0
034	Selenium (mg / kg (ppm))	0939	15.40	1.810	14.40	1.907	0.4013	27	0.52	034.53	0
034	Selenium (mg / kg (ppm))	0186	15.40	0.2000	14.40	1.907	0.4013	27	0.52	034.52	0
034	Selenium (mg / kg (ppm))	0047	15.56	0.0600	14.40	1.907	0.4013	27	0.61	034.52	0
034	Selenium (mg / kg (ppm))	0169	15.60	0.0000	14.40	1.907	0.4013	27	0.63	034.04	0
034	Selenium (mg / kg (ppm))	0918	15.72	1.165	14.40	1.907	0.4013	27	0.69	034.53	0
034	Selenium (mg / kg (ppm))	0036	15.95	0.1000	14.40	1.907	0.4013	27	0.81	034.43	0
034	Selenium (mg / kg (ppm))	0560	15.98	1.110	14.40	1.907	0.4013	27	0.83	034.53	0
034	Selenium (mg / kg (ppm))	0171	16.18	0.2600	14.40	1.907	0.4013	27	0.93	034.04	0
034	Selenium (mg / kg (ppm))	0407	16.97	0.1850	14.40	1.907	0.4013	27	1.35	034.43	0
034	Selenium (mg / kg (ppm))	0692	20.00	0.0000	14.40	1.907	0.4013	27	2.94	034.42	0
034	Selenium (mg / kg (ppm))	0910	14.72	2.410	14.40	1.907	0.4013	27	0.16	034.52	1
035	Sodium (%)	2192	0.3050	0.0100	0.4325	0.0290	0.0128	92	-4.40	035.43	0
035	Sodium (%)	0874	0.3720	0.0020	0.4325	0.0290	0.0128	92	-2.09	035.31	0
035	Sodium (%)	0536	0.3750	0.0100	0.4325	0.0290	0.0128	92	-1.98	035.31	0
035	Sodium (%)	0720	0.3750	0.0100	0.4325	0.0290	0.0128	92	-1.98	035.41	0
035	Sodium (%)	0169	0.3785	0.0170	0.4325	0.0290	0.0128	92	-1.86	035.32	0
035	Sodium (%)	0918	0.3790	0.0300	0.4325	0.0290	0.0128	92	-1.85	035.43	0
035	Sodium (%)	0035	0.3900	0.0000	0.4325	0.0290	0.0128	92	-1.47	035.42	0
035	Sodium (%)	0175	0.3950	0.0300	0.4325	0.0290	0.0128	92	-1.29	035.31	0
035	Sodium (%)	0035	0.3950	0.0100	0.4325	0.0290	0.0128	92	-1.29	035.43	0
035	Sodium (%)	0045	0.3985	0.0270	0.4325	0.0290	0.0128	92	-1.17	035.42	0
035	Sodium (%)	0144	0.4000	0.0200	0.4325	0.0290	0.0128	92	-1.12	035.41	0
035	Sodium (%)	0629	0.4050	0.0100	0.4325	0.0290	0.0128	92	-0.95	035.31	0
035	Sodium (%)	0656	0.4050	0.0100	0.4325	0.0290	0.0128	92	-0.95	035.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
035	Sodium (%)	0300	0.4050	0.0300	0.4325	0.0290	0.0128	92	-0.95	035.99	0
035	Sodium (%)	2129	0.4050	0.0020	0.4325	0.0290	0.0128	92	-0.95	035.99	0
035	Sodium (%)	0868	0.4080	0.0080	0.4325	0.0290	0.0128	92	-0.85	035.01	0
035	Sodium (%)	2006	0.4090	0.0020	0.4325	0.0290	0.0128	92	-0.81	035.01	0
035	Sodium (%)	0510	0.4090	0.0040	0.4325	0.0290	0.0128	92	-0.81	035.43	0
035	Sodium (%)	0358	0.4100	0.0200	0.4325	0.0290	0.0128	92	-0.78	035.41	0
035	Sodium (%)	0520	0.4106	0.0004	0.4325	0.0290	0.0128	92	-0.76	035.41	0
035	Sodium (%)	0619	0.4125	0.0090	0.4325	0.0290	0.0128	92	-0.69	035.41	0
035	Sodium (%)	0042	0.4130	0.0320	0.4325	0.0290	0.0128	92	-0.67	035.42	0
035	Sodium (%)	0553	0.4135	0.0250	0.4325	0.0290	0.0128	92	-0.66	035.41	0
035	Sodium (%)	0884	0.4150	0.0100	0.4325	0.0290	0.0128	92	-0.60	035.02	0
035	Sodium (%)	0019	0.4150	0.0100	0.4325	0.0290	0.0128	92	-0.60	035.41	0
035	Sodium (%)	0511	0.4150	0.0100	0.4325	0.0290	0.0128	92	-0.60	035.41	0
035	Sodium (%)	0205	0.4150	0.0040	0.4325	0.0290	0.0128	92	-0.60	035.42	0
035	Sodium (%)	0278	0.4150	0.0100	0.4325	0.0290	0.0128	92	-0.60	035.42	0
035	Sodium (%)	0098	0.4150	0.0100	0.4325	0.0290	0.0128	92	-0.60	035.43	0
035	Sodium (%)	0505	0.4150	0.0100	0.4325	0.0290	0.0128	92	-0.60	035.43	0
035	Sodium (%)	0870	0.4157	0.0071	0.4325	0.0290	0.0128	92	-0.58	035.42	0
035	Sodium (%)	0345	0.4160	0.0060	0.4325	0.0290	0.0128	92	-0.57	035.43	0
035	Sodium (%)	2196	0.4200	0.0000	0.4325	0.0290	0.0128	92	-0.43	035.31	0
035	Sodium (%)	0294	0.4200	0.0000	0.4325	0.0290	0.0128	92	-0.43	035.42	0
035	Sodium (%)	0682	0.4200	0.0000	0.4325	0.0290	0.0128	92	-0.43	035.43	0
035	Sodium (%)	2089	0.4200	0.0000	0.4325	0.0290	0.0128	92	-0.43	035.43	0
035	Sodium (%)	0693	0.4245	0.0150	0.4325	0.0290	0.0128	92	-0.28	035.42	0
035	Sodium (%)	0868	0.4250	0.0300	0.4325	0.0290	0.0128	92	-0.26	035.31	0
035	Sodium (%)	0171	0.4250	0.0100	0.4325	0.0290	0.0128	92	-0.26	035.41	0
035	Sodium (%)	0425	0.4250	0.0100	0.4325	0.0290	0.0128	92	-0.26	035.43	0
035	Sodium (%)	2022	0.4270	0.0040	0.4325	0.0290	0.0128	92	-0.19	035.31	0
035	Sodium (%)	0202	0.4270	0.0100	0.4325	0.0290	0.0128	92	-0.19	035.43	0
035	Sodium (%)	0407	0.4271	0.0429	0.4325	0.0290	0.0128	92	-0.19	035.41	0
035	Sodium (%)	0049	0.4300	0.0200	0.4325	0.0290	0.0128	92	-0.09	035.41	0
035	Sodium (%)	0148	0.4300	0.0000	0.4325	0.0290	0.0128	92	-0.09	035.41	0
035	Sodium (%)	0190	0.4300	0.0400	0.4325	0.0290	0.0128	92	-0.09	035.42	0
035	Sodium (%)	0032	0.4300	0.0000	0.4325	0.0290	0.0128	92	-0.09	035.43	0
035	Sodium (%)	2055	0.4312	0.0233	0.4325	0.0290	0.0128	92	-0.05	035.53	0
035	Sodium (%)	0265	0.4350	0.0100	0.4325	0.0290	0.0128	92	0.09	035.42	0
035	Sodium (%)	0692	0.4350	0.0100	0.4325	0.0290	0.0128	92	0.09	035.42	0
035	Sodium (%)	0353	0.4350	0.0100	0.4325	0.0290	0.0128	92	0.09	035.43	0
035	Sodium (%)	0861	0.4360	0.0000	0.4325	0.0290	0.0128	92	0.12	035.43	0
035	Sodium (%)	0560	0.4371	0.0113	0.4325	0.0290	0.0128	92	0.16	035.42	0
035	Sodium (%)	0553	0.4385	0.0090	0.4325	0.0290	0.0128	92	0.21	035.53	0
035	Sodium (%)	0229	0.4400	0.0200	0.4325	0.0290	0.0128	92	0.26	035.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
035	Sodium (%)	0100	0.4400	0.0200	0.4325	0.0290	0.0128	92	0.26	035.99	0
035	Sodium (%)	0142	0.4420	0.0420	0.4325	0.0290	0.0128	92	0.33	035.31	0
035	Sodium (%)	0572	0.4420	0.0200	0.4325	0.0290	0.0128	92	0.33	035.53	0
035	Sodium (%)	0164	0.4425	0.0330	0.4325	0.0290	0.0128	92	0.34	035.41	0
035	Sodium (%)	0598	0.4430	0.0144	0.4325	0.0290	0.0128	92	0.36	035.41	0
035	Sodium (%)	0504	0.4450	0.0100	0.4325	0.0290	0.0128	92	0.43	035.05	0
035	Sodium (%)	0689	0.4450	0.0300	0.4325	0.0290	0.0128	92	0.43	035.31	0
035	Sodium (%)	0152	0.4450	0.0100	0.4325	0.0290	0.0128	92	0.43	035.41	0
035	Sodium (%)	0878	0.4450	0.0100	0.4325	0.0290	0.0128	92	0.43	035.41	0
035	Sodium (%)	0186	0.4465	0.0370	0.4325	0.0290	0.0128	92	0.48	035.42	0
035	Sodium (%)	2215	0.4500	0.0000	0.4325	0.0290	0.0128	92	0.60	035.05	0
035	Sodium (%)	0098	0.4500	0.0000	0.4325	0.0290	0.0128	92	0.60	035.41	0
035	Sodium (%)	0242	0.4500	0.0000	0.4325	0.0290	0.0128	92	0.60	035.99	0
035	Sodium (%)	0263	0.4514	0.0112	0.4325	0.0290	0.0128	92	0.65	035.42	0
035	Sodium (%)	0187	0.4547	0.0035	0.4325	0.0290	0.0128	92	0.76	035.42	0
035	Sodium (%)	0037	0.4550	0.0100	0.4325	0.0290	0.0128	92	0.77	035.05	0
035	Sodium (%)	0726	0.4591	0.0005	0.4325	0.0290	0.0128	92	0.91	035.42	0
035	Sodium (%)	0226	0.4600	0.0400	0.4325	0.0290	0.0128	92	0.95	035.41	0
035	Sodium (%)	0083	0.4600	0.0200	0.4325	0.0290	0.0128	92	0.95	035.43	0
035	Sodium (%)	0610	0.4600	0.0200	0.4325	0.0290	0.0128	92	0.95	035.43	0
035	Sodium (%)	0590	0.4600	0.0200	0.4325	0.0290	0.0128	92	0.95	035.99	0
035	Sodium (%)	0354	0.4600	0.0000	0.4325	0.0290	0.0128	92	0.95	035.31	0
035	Sodium (%)	2012	0.4600	0.0000	0.4325	0.0290	0.0128	92	0.95	035.42	0
035	Sodium (%)	0186	0.4600	0.0120	0.4325	0.0290	0.0128	92	0.95	035.52	0
035	Sodium (%)	0948	0.4620	0.0020	0.4325	0.0290	0.0128	92	1.02	035.31	0
035	Sodium (%)	2053	0.4650	0.0100	0.4325	0.0290	0.0128	92	1.12	035.42	0
035	Sodium (%)	0675	0.4650	0.0100	0.4325	0.0290	0.0128	92	1.12	035.43	0
035	Sodium (%)	2152	0.4700	0.0200	0.4325	0.0290	0.0128	92	1.29	035.41	0
035	Sodium (%)	0653	0.4725	0.0010	0.4325	0.0290	0.0128	92	1.38	035.41	0
035	Sodium (%)	0108	0.4800	0.0000	0.4325	0.0290	0.0128	92	1.64	035.05	0
035	Sodium (%)	2195	0.4800	0.0200	0.4325	0.0290	0.0128	92	1.64	035.31	0
035	Sodium (%)	0089	0.4800	0.0000	0.4325	0.0290	0.0128	92	1.64	035.43	0
035	Sodium (%)	2126	0.4850	0.0100	0.4325	0.0290	0.0128	92	1.81	035.31	0
035	Sodium (%)	0123	0.4850	0.0300	0.4325	0.0290	0.0128	92	1.81	035.41	0
035	Sodium (%)	0154	0.4860	0.0257	0.4325	0.0290	0.0128	92	1.84	035.52	0
035	Sodium (%)	0405	0.5100	0.0000	0.4325	0.0290	0.0128	92	2.67	035.41	0
035	Sodium (%)	0609	0.6150	0.0100	0.4325	0.0290	0.0128	92	6.29	035.32	0
035	Sodium (%)	0910	0.4450	0.0700	0.4325	0.0290	0.0128	92	0.43	035.41	1
035	Sodium (%)	0683	4.050	0.0200	0.4325	0.0290	0.0128	92	124.71	035.31	2
036	Sulfur (%)	0553	0.5495	0.0670	0.9114	0.0915	0.0283	35	-3.96	036.99	0
036	Sulfur (%)	0169	0.7250	0.0300	0.9114	0.0915	0.0283	35	-2.04	036.43	0
036	Sulfur (%)	0229	0.7950	0.0300	0.9114	0.0915	0.0283	35	-1.27	036.04	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
036	Sulfur (%)	2129	0.8060	0.0280	0.9114	0.0915	0.0283	35	-1.15	036.42	0
036	Sulfur (%)	0045	0.8125	0.0530	0.9114	0.0915	0.0283	35	-1.08	036.42	0
036	Sulfur (%)	0357	0.8186	0.0243	0.9114	0.0915	0.0283	35	-1.02	036.42	0
036	Sulfur (%)	0098	0.8287	0.0328	0.9114	0.0915	0.0283	35	-0.90	036.43	0
036	Sulfur (%)	0278	0.8300	0.0400	0.9114	0.0915	0.0283	35	-0.89	036.42	0
036	Sulfur (%)	0035	0.8350	0.0700	0.9114	0.0915	0.0283	35	-0.84	036.43	0
036	Sulfur (%)	0345	0.8465	0.0130	0.9114	0.0915	0.0283	35	-0.71	036.43	0
036	Sulfur (%)	0870	0.8518	0.0357	0.9114	0.0915	0.0283	35	-0.65	036.42	0
036	Sulfur (%)	0186	0.8555	0.0650	0.9114	0.0915	0.0283	35	-0.61	036.42	0
036	Sulfur (%)	0918	0.8565	0.0530	0.9114	0.0915	0.0283	35	-0.60	036.43	0
036	Sulfur (%)	0229	0.8950	0.0100	0.9114	0.0915	0.0283	35	-0.18	036.42	0
036	Sulfur (%)	0187	0.9025	0.0040	0.9114	0.0915	0.0283	35	-0.10	036.42	0
036	Sulfur (%)	0693	0.9180	0.0240	0.9114	0.0915	0.0283	35	0.07	036.42	0
036	Sulfur (%)	0560	0.9203	0.0052	0.9114	0.0915	0.0283	35	0.10	036.42	0
036	Sulfur (%)	0242	0.9250	0.0100	0.9114	0.0915	0.0283	35	0.15	036.99	0
036	Sulfur (%)	0202	0.9255	0.0210	0.9114	0.0915	0.0283	35	0.15	036.43	0
036	Sulfur (%)	0171	0.9320	0.0060	0.9114	0.0915	0.0283	35	0.22	036.42	0
036	Sulfur (%)	0407	0.9351	0.0163	0.9114	0.0915	0.0283	35	0.26	036.42	0
036	Sulfur (%)	0148	0.9376	0.0104	0.9114	0.0915	0.0283	35	0.29	036.04	0
036	Sulfur (%)	0553	0.9395	0.0750	0.9114	0.0915	0.0283	35	0.31	036.53	0
036	Sulfur (%)	0186	0.9470	0.0260	0.9114	0.0915	0.0283	35	0.39	036.52	0
036	Sulfur (%)	0692	0.9500	0.0000	0.9114	0.0915	0.0283	35	0.42	036.42	0
036	Sulfur (%)	0035	0.9600	0.0600	0.9114	0.0915	0.0283	35	0.53	036.42	0
036	Sulfur (%)	0037	0.9900	0.0000	0.9114	0.0915	0.0283	35	0.86	036.42	0
036	Sulfur (%)	0353	0.9900	0.0400	0.9114	0.0915	0.0283	35	0.86	036.43	0
036	Sulfur (%)	0294	1.000	0.0000	0.9114	0.0915	0.0283	35	0.97	036.42	0
036	Sulfur (%)	0032	1.010	0.0000	0.9114	0.0915	0.0283	35	1.08	036.43	0
036	Sulfur (%)	0083	1.010	0.0400	0.9114	0.0915	0.0283	35	1.08	036.43	0
036	Sulfur (%)	2089	1.015	0.0700	0.9114	0.0915	0.0283	35	1.13	036.43	0
036	Sulfur (%)	0265	1.020	0.0000	0.9114	0.0915	0.0283	35	1.19	036.42	0
036	Sulfur (%)	0510	1.045	0.0100	0.9114	0.0915	0.0283	35	1.46	036.43	0
036	Sulfur (%)	0226	1.070	0.0200	0.9114	0.0915	0.0283	35	1.73	036.04	0
036	Sulfur (%)	0910	7,599	151.0	0.9114	0.0915	0.0283	35	83065.62	036.52	2
037	Zinc (mg / kg (ppm))	0202	0.3845	0.0010	3,779	316.4	103.5	97	-11.94	037.43	0
037	Zinc (mg / kg (ppm))	0006	0.3955	0.0030	3,779	316.4	103.5	97	-11.94	037.43	0
037	Zinc (mg / kg (ppm))	0726	4.011	0.0210	3,779	316.4	103.5	97	-11.93	037.42	0
037	Zinc (mg / kg (ppm))	0265	2,530	460.0	3,779	316.4	103.5	97	-3.95	037.42	0
037	Zinc (mg / kg (ppm))	0345	2,760	40.00	3,779	316.4	103.5	97	-3.22	037.43	0
037	Zinc (mg / kg (ppm))	0047	2,806	55.00	3,779	316.4	103.5	97	-3.08	037.52	0
037	Zinc (mg / kg (ppm))	0169	2,970	80.00	3,779	316.4	103.5	97	-2.56	037.43	0
037	Zinc (mg / kg (ppm))	0045	3,030	20.00	3,779	316.4	103.5	97	-2.37	037.42	0
037	Zinc (mg / kg (ppm))	0405	3,122	0.0000	3,779	316.4	103.5	97	-2.08	037.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
037	Zinc (mg / kg (ppm))	0619	3,219	24.00	3,779	316.4	103.5	97	-1.77	037.41	0
037	Zinc (mg / kg (ppm))	2192	3,231	337.2	3,779	316.4	103.5	97	-1.73	037.43	0
037	Zinc (mg / kg (ppm))	0187	3,235	4.080	3,779	316.4	103.5	97	-1.72	037.42	0
037	Zinc (mg / kg (ppm))	0689	3,292	126.3	3,779	316.4	103.5	97	-1.54	037.31	0
037	Zinc (mg / kg (ppm))	2055	3,381	255.0	3,779	316.4	103.5	97	-1.26	037.53	0
037	Zinc (mg / kg (ppm))	0190	3,436	8.000	3,779	316.4	103.5	97	-1.09	037.42	0
037	Zinc (mg / kg (ppm))	0598	3,461	19.00	3,779	316.4	103.5	97	-1.01	037.41	0
037	Zinc (mg / kg (ppm))	0186	3,463	98.00	3,779	316.4	103.5	97	-1.00	037.42	0
037	Zinc (mg / kg (ppm))	2128	3,497	148.6	3,779	316.4	103.5	97	-0.89	037.32	0
037	Zinc (mg / kg (ppm))	0049	3,501	116.7	3,779	316.4	103.5	97	-0.88	037.41	0
037	Zinc (mg / kg (ppm))	0037	3,512	136.0	3,779	316.4	103.5	97	-0.85	037.42	0
037	Zinc (mg / kg (ppm))	0164	3,555	30.00	3,779	316.4	103.5	97	-0.71	037.44	0
037	Zinc (mg / kg (ppm))	0553	3,588	148.0	3,779	316.4	103.5	97	-0.61	037.41	0
037	Zinc (mg / kg (ppm))	0610	3,609	35.00	3,779	316.4	103.5	97	-0.54	037.43	0
037	Zinc (mg / kg (ppm))	0511	3,617	23.00	3,779	316.4	103.5	97	-0.52	037.41	0
037	Zinc (mg / kg (ppm))	0263	3,618	45.98	3,779	316.4	103.5	97	-0.51	037.42	0
037	Zinc (mg / kg (ppm))	0357	3,628	28.48	3,779	316.4	103.5	97	-0.48	037.42	0
037	Zinc (mg / kg (ppm))	0032	3,650	40.00	3,779	316.4	103.5	97	-0.41	037.43	0
037	Zinc (mg / kg (ppm))	0505	3,654	29.00	3,779	316.4	103.5	97	-0.40	037.43	0
037	Zinc (mg / kg (ppm))	0536	3,661	57.50	3,779	316.4	103.5	97	-0.38	037.31	0
037	Zinc (mg / kg (ppm))	0098	3,661	64.00	3,779	316.4	103.5	97	-0.37	037.43	0
037	Zinc (mg / kg (ppm))	0242	3,664	208.0	3,779	316.4	103.5	97	-0.36	037.99	0
037	Zinc (mg / kg (ppm))	0294	3,682	6.500	3,779	316.4	103.5	97	-0.31	037.42	0
037	Zinc (mg / kg (ppm))	0098	3,690	115.0	3,779	316.4	103.5	97	-0.28	037.41	0
037	Zinc (mg / kg (ppm))	2062	3,692	3.800	3,779	316.4	103.5	97	-0.28	037.31	0
037	Zinc (mg / kg (ppm))	0001	3,693	257.0	3,779	316.4	103.5	97	-0.27	037.31	0
037	Zinc (mg / kg (ppm))	0027	3,693	50.60	3,779	316.4	103.5	97	-0.27	037.43	0
037	Zinc (mg / kg (ppm))	0520	3,707	170.0	3,779	316.4	103.5	97	-0.23	037.41	0
037	Zinc (mg / kg (ppm))	0009	3,713	303.0	3,779	316.4	103.5	97	-0.21	037.42	0
037	Zinc (mg / kg (ppm))	0042	3,713	303.0	3,779	316.4	103.5	97	-0.21	037.42	0
037	Zinc (mg / kg (ppm))	0964	3,725	72.98	3,779	316.4	103.5	97	-0.17	037.43	0
037	Zinc (mg / kg (ppm))	2022	3,741	0.3220	3,779	316.4	103.5	97	-0.12	037.31	0
037	Zinc (mg / kg (ppm))	0560	3,743	32.60	3,779	316.4	103.5	97	-0.12	037.42	0
037	Zinc (mg / kg (ppm))	0653	3,751	38.00	3,779	316.4	103.5	97	-0.09	037.41	0
037	Zinc (mg / kg (ppm))	0693	3,760	161.7	3,779	316.4	103.5	97	-0.06	037.42	0
037	Zinc (mg / kg (ppm))	0035	3,767	213.0	3,779	316.4	103.5	97	-0.04	037.43	0
037	Zinc (mg / kg (ppm))	0010	3,770	40.00	3,779	316.4	103.5	97	-0.03	037.33	0
037	Zinc (mg / kg (ppm))	0510	3,784	8.000	3,779	316.4	103.5	97	0.01	037.43	0
037	Zinc (mg / kg (ppm))	0033	3,795	110.0	3,779	316.4	103.5	97	0.05	037.43	0
037	Zinc (mg / kg (ppm))	0572	3,795	290.0	3,779	316.4	103.5	97	0.05	037.53	0
037	Zinc (mg / kg (ppm))	0682	3,801	0.0000	3,779	316.4	103.5	97	0.07	037.43	0
037	Zinc (mg / kg (ppm))	0019	3,807	144.0	3,779	316.4	103.5	97	0.09	037.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
037	Zinc (mg / kg (ppm))	0407	3,809	127.2	3,779	316.4	103.5	97	0.09	037.41	0
037	Zinc (mg / kg (ppm))	0016	3,815	30.00	3,779	316.4	103.5	97	0.11	037.52	0
037	Zinc (mg / kg (ppm))	0278	3,835	67.70	3,779	316.4	103.5	97	0.18	037.42	0
037	Zinc (mg / kg (ppm))	0870	3,838	209.0	3,779	316.4	103.5	97	0.18	037.42	0
037	Zinc (mg / kg (ppm))	0861	3,840	120.0	3,779	316.4	103.5	97	0.19	037.43	0
037	Zinc (mg / kg (ppm))	2089	3,841	95.18	3,779	316.4	103.5	97	0.19	037.43	0
037	Zinc (mg / kg (ppm))	0148	3,844	86.22	3,779	316.4	103.5	97	0.21	037.41	0
037	Zinc (mg / kg (ppm))	0074	3,849	25.00	3,779	316.4	103.5	97	0.22	037.41	0
037	Zinc (mg / kg (ppm))	2144	3,860	26.72	3,779	316.4	103.5	97	0.25	037.34	0
037	Zinc (mg / kg (ppm))	2196	3,873	0.0000	3,779	316.4	103.5	97	0.30	037.31	0
037	Zinc (mg / kg (ppm))	0910	3,877	326.0	3,779	316.4	103.5	97	0.31	037.41	0
037	Zinc (mg / kg (ppm))	2188	3,885	98.41	3,779	316.4	103.5	97	0.33	037.31	0
037	Zinc (mg / kg (ppm))	0035	3,904	104.0	3,779	316.4	103.5	97	0.39	037.42	0
037	Zinc (mg / kg (ppm))	0723	3,912	0.0000	3,779	316.4	103.5	97	0.42	037.31	0
037	Zinc (mg / kg (ppm))	0590	3,918	11.00	3,779	316.4	103.5	97	0.44	037.99	0
037	Zinc (mg / kg (ppm))	0868	3,920	76.90	3,779	316.4	103.5	97	0.44	037.31	0
037	Zinc (mg / kg (ppm))	0003	3,920	292.0	3,779	316.4	103.5	97	0.44	037.41	0
037	Zinc (mg / kg (ppm))	0656	3,927	10.64	3,779	316.4	103.5	97	0.47	037.31	0
037	Zinc (mg / kg (ppm))	0939	3,941	42.00	3,779	316.4	103.5	97	0.51	037.53	0
037	Zinc (mg / kg (ppm))	0175	3,950	100.0	3,779	316.4	103.5	97	0.54	037.31	0
037	Zinc (mg / kg (ppm))	0083	3,951	107.0	3,779	316.4	103.5	97	0.54	037.43	0
037	Zinc (mg / kg (ppm))	0848	3,958	8.930	3,779	316.4	103.5	97	0.57	037.41	0
037	Zinc (mg / kg (ppm))	0229	3,987	23.00	3,779	316.4	103.5	97	0.65	037.41	0
037	Zinc (mg / kg (ppm))	0186	3,990	40.00	3,779	316.4	103.5	97	0.67	037.52	0
037	Zinc (mg / kg (ppm))	0948	3,996	73.38	3,779	316.4	103.5	97	0.69	037.33	0
037	Zinc (mg / kg (ppm))	0553	4,010	340.0	3,779	316.4	103.5	97	0.73	037.53	0
037	Zinc (mg / kg (ppm))	0918	4,049	416.6	3,779	316.4	103.5	97	0.85	037.43	0
037	Zinc (mg / kg (ppm))	2195	4,054	15.47	3,779	316.4	103.5	97	0.87	037.31	0
037	Zinc (mg / kg (ppm))	0425	4,061	42.50	3,779	316.4	103.5	97	0.89	037.43	0
037	Zinc (mg / kg (ppm))	0874	4,080	160.0	3,779	316.4	103.5	97	0.95	037.31	0
037	Zinc (mg / kg (ppm))	0226	4,088	64.00	3,779	316.4	103.5	97	0.98	037.41	0
037	Zinc (mg / kg (ppm))	2012	4,088	360.8	3,779	316.4	103.5	97	0.98	037.42	0
037	Zinc (mg / kg (ppm))	2145	4,107	12.70	3,779	316.4	103.5	97	1.04	037.41	0
037	Zinc (mg / kg (ppm))	0878	4,120	30.00	3,779	316.4	103.5	97	1.08	037.41	0
037	Zinc (mg / kg (ppm))	2129	4,133	695.0	3,779	316.4	103.5	97	1.12	037.99	0
037	Zinc (mg / kg (ppm))	2146	4,157	27.00	3,779	316.4	103.5	97	1.19	037.51	0
037	Zinc (mg / kg (ppm))	0354	4,164	17.81	3,779	316.4	103.5	97	1.21	037.31	0
037	Zinc (mg / kg (ppm))	2178	4,180	34.00	3,779	316.4	103.5	97	1.27	037.33	0
037	Zinc (mg / kg (ppm))	0720	4,187	193.3	3,779	316.4	103.5	97	1.29	037.41	0
037	Zinc (mg / kg (ppm))	0028	4,240	72.17	3,779	316.4	103.5	97	1.45	037.43	0
037	Zinc (mg / kg (ppm))	0968	4,249	6.000	3,779	316.4	103.5	97	1.48	037.43	0
037	Zinc (mg / kg (ppm))	0353	4,312	18.00	3,779	316.4	103.5	97	1.68	037.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
037	Zinc (mg / kg (ppm))	0171	4,351	39.00	3,779	316.4	103.5	97	1.80	037.41	0
037	Zinc (mg / kg (ppm))	0358	4,428	134.0	3,779	316.4	103.5	97	2.05	037.41	0
037	Zinc (mg / kg (ppm))	0100	4,510	289.0	3,779	316.4	103.5	97	2.31	037.99	0
037	Zinc (mg / kg (ppm))	0675	5,177	12.68	3,779	316.4	103.5	97	4.42	037.43	0
037	Zinc (mg / kg (ppm))	0692	5,055	2,030	3,779	316.4	103.5	97	4.03	037.42	1
038	Molybdenum (mg / kg (ppm))	0169	4.295	0.0100	5.438	0.8112	0.2137	20	-1.41	038.43	0
038	Molybdenum (mg / kg (ppm))	0345	4.445	0.0900	5.438	0.8112	0.2137	20	-1.22	038.43	0
038	Molybdenum (mg / kg (ppm))	0226	4.505	0.4500	5.438	0.8112	0.2137	20	-1.15	038.41	0
038	Molybdenum (mg / kg (ppm))	0353	4.610	0.0501	5.438	0.8112	0.2137	20	-1.02	038.43	0
038	Molybdenum (mg / kg (ppm))	0560	4.765	0.1500	5.438	0.8112	0.2137	20	-0.83	038.42	0
038	Molybdenum (mg / kg (ppm))	2055	4.900	0.2800	5.438	0.8112	0.2137	20	-0.66	038.53	0
038	Molybdenum (mg / kg (ppm))	0265	4.935	0.2900	5.438	0.8112	0.2137	20	-0.62	038.42	0
038	Molybdenum (mg / kg (ppm))	0278	5.135	0.2700	5.438	0.8112	0.2137	20	-0.37	038.42	0
038	Molybdenum (mg / kg (ppm))	0910	5.265	0.2500	5.438	0.8112	0.2137	20	-0.21	038.52	0
038	Molybdenum (mg / kg (ppm))	0693	5.356	0.2430	5.438	0.8112	0.2137	20	-0.10	038.42	0
038	Molybdenum (mg / kg (ppm))	0918	5.695	0.0586	5.438	0.8112	0.2137	20	0.32	038.53	0
038	Molybdenum (mg / kg (ppm))	0148	5.724	0.0530	5.438	0.8112	0.2137	20	0.35	038.41	0
038	Molybdenum (mg / kg (ppm))	0164	5.725	0.2100	5.438	0.8112	0.2137	20	0.35	038.53	0
038	Molybdenum (mg / kg (ppm))	0510	5.800	0.0000	5.438	0.8112	0.2137	20	0.45	038.43	0
038	Molybdenum (mg / kg (ppm))	0186	6.000	0.0000	5.438	0.8112	0.2137	20	0.69	038.52	0
038	Molybdenum (mg / kg (ppm))	0553	6.120	0.1000	5.438	0.8112	0.2137	20	0.84	038.53	0
038	Molybdenum (mg / kg (ppm))	0045	6.150	0.9000	5.438	0.8112	0.2137	20	0.88	038.42	0
038	Molybdenum (mg / kg (ppm))	0083	6.300	0.2000	5.438	0.8112	0.2137	20	1.06	038.43	0
038	Molybdenum (mg / kg (ppm))	2053	6.375	0.4500	5.438	0.8112	0.2137	20	1.16	038.42	0
038	Molybdenum (mg / kg (ppm))	0032	6.960	0.2200	5.438	0.8112	0.2137	20	1.88	038.43	0
038	Molybdenum (mg / kg (ppm))	0098	5.081	1.703	5.438	0.8112	0.2137	20	-0.44	038.53	1
040	Barium (mg / kg (ppm))	0918	34.98	0.5849				2		040.53	0
040	Barium (mg / kg (ppm))	0560	38.00	0.4000				2		040.42	0
041	Vanadium (mg / kg (ppm))	0553	12.00	0.6000				1		041.53	0
042	Chloride (%)	0948	1.200	0.0200				1		042.00	0
099	Menadione (form) (mg / kg (ppm))	2192	26.30	1.200				1		099.01	0
101	Choline Chloride (mg / kg (ppm))	0939	4,173	21.00				2		101.99	0
101	Choline Chloride (mg / kg (ppm))	0160	4,665	250.0				2		101.99	0
102	Niacin (mg / kg (ppm))	2022	198.5	21.00	2,003	416.3	69.08	6	-4.33	102.02	0
102	Niacin (mg / kg (ppm))	2146	1,860	20.00	2,003	416.3	69.08	6	-0.34	102.99	0
102	Niacin (mg / kg (ppm))	0227	1,965	10.00	2,003	416.3	69.08	6	-0.09	102.01	0
102	Niacin (mg / kg (ppm))	0939	2,096	70.00	2,003	416.3	69.08	6	0.22	102.02	0
102	Niacin (mg / kg (ppm))	0010	2,350	180.0	2,003	416.3	69.08	6	0.83	102.99	0
102	Niacin (mg / kg (ppm))	0904	2,367	113.5	2,003	416.3	69.08	6	0.88	102.99	0
103	Pantothenic Acid (mg / kg (ppm))	0939	203.3	1.700				3		103.02	0
103	Pantothenic Acid (mg / kg (ppm))	2103	252.5	5.000				3		103.02	0
103	Pantothenic Acid (mg / kg (ppm))	2146	402.2	4.900				3		103.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
103	Pantothenic Acid (mg / kg (ppm))	0010	363.5	39.60				3		103.99	1
104	Riboflavin (mg / kg (ppm))	0910	603.3	76.73	831.9	58.45	30.59	10	-3.91	104.03	0
104	Riboflavin (mg / kg (ppm))	2103	787.5	25.00	831.9	58.45	30.59	10	-0.76	104.03	0
104	Riboflavin (mg / kg (ppm))	0939	799.0	5.170	831.9	58.45	30.59	10	-0.56	104.03	0
104	Riboflavin (mg / kg (ppm))	0227	801.5	29.00	831.9	58.45	30.59	10	-0.52	104.00	0
104	Riboflavin (mg / kg (ppm))	0848	826.0	42.00	831.9	58.45	30.59	10	-0.10	104.03	0
104	Riboflavin (mg / kg (ppm))	2022	833.0	44.00	831.9	58.45	30.59	10	0.02	104.03	0
104	Riboflavin (mg / kg (ppm))	0171	861.0	28.00	831.9	58.45	30.59	10	0.50	104.00	0
104	Riboflavin (mg / kg (ppm))	0904	863.9	52.17	831.9	58.45	30.59	10	0.55	104.99	0
104	Riboflavin (mg / kg (ppm))	2152	883.5	1.940	831.9	58.45	30.59	10	0.88	104.03	0
104	Riboflavin (mg / kg (ppm))	2146	920.3	1.900	831.9	58.45	30.59	10	1.51	104.99	0
104	Riboflavin (mg / kg (ppm))	0010	921.5	197.0	831.9	58.45	30.59	10	1.53	104.99	1
105	Thiamine (mg / kg (ppm))	2103	277.0	36.00	332.2	50.52	32.16	9	-1.09	105.00	0
105	Thiamine (mg / kg (ppm))	0939	277.1	3.590	332.2	50.52	32.16	9	-1.09	105.00	0
105	Thiamine (mg / kg (ppm))	0227	307.0	44.00	332.2	50.52	32.16	9	-0.50	105.01	0
105	Thiamine (mg / kg (ppm))	0160	309.0	46.00	332.2	50.52	32.16	9	-0.46	105.01	0
105	Thiamine (mg / kg (ppm))	2146	312.6	4.900	332.2	50.52	32.16	9	-0.39	105.99	0
105	Thiamine (mg / kg (ppm))	0904	364.2	20.80	332.2	50.52	32.16	9	0.63	105.99	0
105	Thiamine (mg / kg (ppm))	2053	377.5	37.00	332.2	50.52	32.16	9	0.90	105.00	0
105	Thiamine (mg / kg (ppm))	2022	380.5	9.000	332.2	50.52	32.16	9	0.96	105.00	0
105	Thiamine (mg / kg (ppm))	0910	385.4	88.16	332.2	50.52	32.16	9	1.05	105.00	0
106	Vitamin A (KU / kg)	0905	63.75	12.24	81.17	12.36	5.083	22	-1.41	106.02	0
106	Vitamin A (KU / kg)	2195	64.10	0.0000	81.17	12.36	5.083	22	-1.38	106.02	0
106	Vitamin A (KU / kg)	2022	66.50	9.000	81.17	12.36	5.083	22	-1.19	106.02	0
106	Vitamin A (KU / kg)	0610	67.90	5.400	81.17	12.36	5.083	22	-1.07	106.02	0
106	Vitamin A (KU / kg)	0689	75.45	1.700	81.17	12.36	5.083	22	-0.46	106.02	0
106	Vitamin A (KU / kg)	0098	75.90	10.20	81.17	12.36	5.083	22	-0.43	106.01	0
106	Vitamin A (KU / kg)	0227	76.05	6.500	81.17	12.36	5.083	22	-0.41	106.02	0
106	Vitamin A (KU / kg)	0904	76.28	5.830	81.17	12.36	5.083	22	-0.40	106.99	0
106	Vitamin A (KU / kg)	0035	76.82	0.0100	81.17	12.36	5.083	22	-0.35	106.02	0
106	Vitamin A (KU / kg)	0910	77.60	2.600	81.17	12.36	5.083	22	-0.29	106.02	0
106	Vitamin A (KU / kg)	2053	77.60	5.000	81.17	12.36	5.083	22	-0.29	106.02	0
106	Vitamin A (KU / kg)	2103	79.25	5.010	81.17	12.36	5.083	22	-0.16	106.02	0
106	Vitamin A (KU / kg)	0861	81.10	1.295	81.17	12.36	5.083	22	-0.01	106.02	0
106	Vitamin A (KU / kg)	0148	83.51	8.910	81.17	12.36	5.083	22	0.19	106.02	0
106	Vitamin A (KU / kg)	2192	85.94	4.779	81.17	12.36	5.083	22	0.39	106.02	0
106	Vitamin A (KU / kg)	0875	88.73	1.280	81.17	12.36	5.083	22	0.61	106.02	0
106	Vitamin A (KU / kg)	0676	88.77	2.800	81.17	12.36	5.083	22	0.61	106.02	0
106	Vitamin A (KU / kg)	0010	88.85	20.30	81.17	12.36	5.083	22	0.62	106.02	0
106	Vitamin A (KU / kg)	0004	93.50	5.000	81.17	12.36	5.083	22	1.00	106.02	0
106	Vitamin A (KU / kg)	0171	98.75	0.5000	81.17	12.36	5.083	22	1.42	106.00	0
106	Vitamin A (KU / kg)	0019	331.3	0.8010	81.17	12.36	5.083	22	20.23	106.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
106	Vitamin A (KU / kg)	0675	407.6	2.680	81.17	12.36	5.083	22	26.40	106.02	0
107	Vitamin B12 (µg / kg (ppb))	0227	1,070	0.0000				1		107.00	0
108	Vitamin D3 (KU / kg)	0610	8.650	1.300	28.81	17.29	0.7200	7	-1.17	108.02	0
108	Vitamin D3 (KU / kg)	0861	18.54	1.480	28.81	17.29	0.7200	7	-0.59	108.02	0
108	Vitamin D3 (KU / kg)	2192	24.25	0.4200	28.81	17.29	0.7200	7	-0.26	108.02	0
108	Vitamin D3 (KU / kg)	0910	25.40	0.6700	28.81	17.29	0.7200	7	-0.20	108.02	0
108	Vitamin D3 (KU / kg)	0227	27.95	0.7000	28.81	17.29	0.7200	7	-0.05	108.02	0
108	Vitamin D3 (KU / kg)	0675	42.13	0.4700	28.81	17.29	0.7200	7	0.77	108.02	0
108	Vitamin D3 (KU / kg)	2195	387.8	0.0000	28.81	17.29	0.7200	7	20.76	108.02	0
109	Vitamin E (IU / kg)	0905	1,192	226.0	1,636	335.3	72.68	17	-1.32	109.02	0
109	Vitamin E (IU / kg)	0148	1,239	319.5	1,636	335.3	72.68	17	-1.18	109.02	0
109	Vitamin E (IU / kg)	0875	1,270	57.12	1,636	335.3	72.68	17	-1.09	109.02	0
109	Vitamin E (IU / kg)	0610	1,380	40.00	1,636	335.3	72.68	17	-0.76	109.02	0
109	Vitamin E (IU / kg)	2195	1,393	0.0000	1,636	335.3	72.68	17	-0.72	109.02	0
109	Vitamin E (IU / kg)	0171	1,500	0.0000	1,636	335.3	72.68	17	-0.40	109.99	0
109	Vitamin E (IU / kg)	0861	1,580	40.00	1,636	335.3	72.68	17	-0.17	109.02	0
109	Vitamin E (IU / kg)	2103	1,607	10.00	1,636	335.3	72.68	17	-0.09	109.02	0
109	Vitamin E (IU / kg)	0910	1,626	63.10	1,636	335.3	72.68	17	-0.03	109.02	0
109	Vitamin E (IU / kg)	0675	1,648	16.66	1,636	335.3	72.68	17	0.04	109.02	0
109	Vitamin E (IU / kg)	2053	1,648	98.00	1,636	335.3	72.68	17	0.04	109.02	0
109	Vitamin E (IU / kg)	0098	1,740	20.00	1,636	335.3	72.68	17	0.31	109.02	0
109	Vitamin E (IU / kg)	0227	1,755	270.0	1,636	335.3	72.68	17	0.36	109.02	0
109	Vitamin E (IU / kg)	0904	1,952	6.260	1,636	335.3	72.68	17	0.94	109.99	0
109	Vitamin E (IU / kg)	2022	2,000	0.0000	1,636	335.3	72.68	17	1.09	109.02	0
109	Vitamin E (IU / kg)	0676	2,174	49.00	1,636	335.3	72.68	17	1.60	109.02	0
109	Vitamin E (IU / kg)	2192	2,355	20.00	1,636	335.3	72.68	17	2.14	109.02	0
112	Pyridoxine (µg / g)	2146	220.5	7.400	248.6	19.63	16.79	8	-1.43	112.99	0
112	Pyridoxine (µg / g)	0904	232.4	16.55	248.6	19.63	16.79	8	-0.82	112.99	0
112	Pyridoxine (µg / g)	0010	238.0	42.00	248.6	19.63	16.79	8	-0.54	112.99	0
112	Pyridoxine (µg / g)	2022	247.0	40.00	248.6	19.63	16.79	8	-0.08	112.01	0
112	Pyridoxine (µg / g)	2103	254.5	1.000	248.6	19.63	16.79	8	0.30	112.01	0
112	Pyridoxine (µg / g)	2152	260.6	0.6400	248.6	19.63	16.79	8	0.61	112.01	0
112	Pyridoxine (µg / g)	0939	265.8	12.70	248.6	19.63	16.79	8	0.87	112.01	0
112	Pyridoxine (µg / g)	0227	270.0	14.00	248.6	19.63	16.79	8	1.09	112.01	0
113	Folic Acid (mg / kg (ppm))	2146	18.45	0.3100				3		113.99	0
113	Folic Acid (mg / kg (ppm))	0227	26.20	2.200				3		113.01	0
113	Folic Acid (mg / kg (ppm))	2022	157.0	4.000				3		113.02	0
114	Biotin (mg / kg (ppm))	0227	0.7720	0.0680				1		114.01	0
118	Peroxide value (meq/kg)	2081	2.105	1.090				1		118.99	0
120	Alanine (%)	2196	0.5260	0.0000	0.6118	0.0370	0.0071	29	-2.32	120.05	0
120	Alanine (%)	0859	0.5300	0.0120	0.6118	0.0370	0.0071	29	-2.21	120.00	0
120	Alanine (%)	0870	0.5357	0.0178	0.6118	0.0370	0.0071	29	-2.06	120.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
120	Alanine (%)	0354	0.5590	0.0060	0.6118	0.0370	0.0071	29	-1.43	120.00	0
120	Alanine (%)	0910	0.5650	0.0100	0.6118	0.0370	0.0071	29	-1.27	120.00	0
120	Alanine (%)	0407	0.5755	0.0110	0.6118	0.0370	0.0071	29	-0.98	120.05	0
120	Alanine (%)	0148	0.5850	0.0100	0.6118	0.0370	0.0071	29	-0.72	120.05	0
120	Alanine (%)	0868	0.6030	0.0040	0.6118	0.0370	0.0071	29	-0.24	120.00	0
120	Alanine (%)	0968	0.6030	0.0100	0.6118	0.0370	0.0071	29	-0.24	120.00	0
120	Alanine (%)	0878	0.6060	0.0080	0.6118	0.0370	0.0071	29	-0.16	120.00	0
120	Alanine (%)	0918	0.6063	0.0034	0.6118	0.0370	0.0071	29	-0.15	120.00	0
120	Alanine (%)	0675	0.6100	0.0200	0.6118	0.0370	0.0071	29	-0.05	120.00	0
120	Alanine (%)	0610	0.6100	0.0000	0.6118	0.0370	0.0071	29	-0.05	120.05	0
120	Alanine (%)	0676	0.6100	0.0200	0.6118	0.0370	0.0071	29	-0.05	120.05	0
120	Alanine (%)	2195	0.6100	0.0000	0.6118	0.0370	0.0071	29	-0.05	120.05	0
120	Alanine (%)	0644	0.6110	0.0040	0.6118	0.0370	0.0071	29	-0.02	120.00	0
120	Alanine (%)	0504	0.6200	0.0000	0.6118	0.0370	0.0071	29	0.22	120.00	0
120	Alanine (%)	0571	0.6225	0.0030	0.6118	0.0370	0.0071	29	0.29	120.00	0
120	Alanine (%)	0227	0.6250	0.0100	0.6118	0.0370	0.0071	29	0.36	120.00	0
120	Alanine (%)	2059	0.6270	0.0020	0.6118	0.0370	0.0071	29	0.41	120.00	0
120	Alanine (%)	0872	0.6275	0.0050	0.6118	0.0370	0.0071	29	0.42	120.00	0
120	Alanine (%)	0098	0.6300	0.0060	0.6118	0.0370	0.0071	29	0.49	120.02	0
120	Alanine (%)	0353	0.6400	0.0000	0.6118	0.0370	0.0071	29	0.76	120.00	0
120	Alanine (%)	0619	0.6445	0.0110	0.6118	0.0370	0.0071	29	0.88	120.00	0
120	Alanine (%)	0626	0.6490	0.0080	0.6118	0.0370	0.0071	29	1.01	120.05	0
120	Alanine (%)	0171	0.6515	0.0130	0.6118	0.0370	0.0071	29	1.07	120.00	0
120	Alanine (%)	0226	0.6546	0.0027	0.6118	0.0370	0.0071	29	1.16	120.00	0
120	Alanine (%)	0682	0.6610	0.0000	0.6118	0.0370	0.0071	29	1.33	120.00	0
120	Alanine (%)	2188	0.7350	0.0100	0.6118	0.0370	0.0071	29	3.33	120.05	0
120	Alanine (%)	0853	0.7000	0.1000	0.6118	0.0370	0.0071	29	2.38	120.05	1
121	Arginine (%)	0626	0.5965	0.0070	0.7306	0.0439	0.0141	30	-3.05	121.05	0
121	Arginine (%)	0859	0.6025	0.0082	0.7306	0.0439	0.0141	30	-2.92	121.00	0
121	Arginine (%)	0870	0.6459	0.0028	0.7306	0.0439	0.0141	30	-1.93	121.00	0
121	Arginine (%)	0354	0.6930	0.0000	0.7306	0.0439	0.0141	30	-0.86	121.00	0
121	Arginine (%)	0407	0.6945	0.0130	0.7306	0.0439	0.0141	30	-0.82	121.05	0
121	Arginine (%)	0226	0.6947	0.0430	0.7306	0.0439	0.0141	30	-0.82	121.00	0
121	Arginine (%)	0353	0.7000	0.0200	0.7306	0.0439	0.0141	30	-0.70	121.00	0
121	Arginine (%)	0171	0.7010	0.0240	0.7306	0.0439	0.0141	30	-0.67	121.00	0
121	Arginine (%)	0968	0.7095	0.0010	0.7306	0.0439	0.0141	30	-0.48	121.00	0
121	Arginine (%)	0675	0.7100	0.0200	0.7306	0.0439	0.0141	30	-0.47	121.00	0
121	Arginine (%)	0918	0.7130	0.0034	0.7306	0.0439	0.0141	30	-0.40	121.00	0
121	Arginine (%)	0504	0.7150	0.0100	0.7306	0.0439	0.0141	30	-0.36	121.00	0
121	Arginine (%)	0227	0.7200	0.0200	0.7306	0.0439	0.0141	30	-0.24	121.00	0
121	Arginine (%)	0682	0.7200	0.0000	0.7306	0.0439	0.0141	30	-0.24	121.00	0
121	Arginine (%)	0098	0.7215	0.0030	0.7306	0.0439	0.0141	30	-0.21	121.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
121	Arginine (%)	0878	0.7225	0.0090	0.7306	0.0439	0.0141	30	-0.18	121.00	0
121	Arginine (%)	0853	0.7300	0.0600	0.7306	0.0439	0.0141	30	-0.01	121.05	0
121	Arginine (%)	0910	0.7450	0.0100	0.7306	0.0439	0.0141	30	0.33	121.00	0
121	Arginine (%)	0571	0.7490	0.0060	0.7306	0.0439	0.0141	30	0.42	121.00	0
121	Arginine (%)	0148	0.7500	0.0400	0.7306	0.0439	0.0141	30	0.44	121.05	0
121	Arginine (%)	0872	0.7510	0.0100	0.7306	0.0439	0.0141	30	0.46	121.00	0
121	Arginine (%)	2059	0.7520	0.0020	0.7306	0.0439	0.0141	30	0.49	121.00	0
121	Arginine (%)	0610	0.7550	0.0100	0.7306	0.0439	0.0141	30	0.56	121.05	0
121	Arginine (%)	0644	0.7580	0.0080	0.7306	0.0439	0.0141	30	0.62	121.00	0
121	Arginine (%)	0868	0.7630	0.0140	0.7306	0.0439	0.0141	30	0.74	121.00	0
121	Arginine (%)	2196	0.7770	0.0000	0.7306	0.0439	0.0141	30	1.06	121.05	0
121	Arginine (%)	0676	0.7900	0.0400	0.7306	0.0439	0.0141	30	1.35	121.05	0
121	Arginine (%)	2188	0.8050	0.0300	0.7306	0.0439	0.0141	30	1.69	121.05	0
121	Arginine (%)	2195	0.8300	0.0000	0.7306	0.0439	0.0141	30	2.26	121.05	0
121	Arginine (%)	0619	0.9940	0.0100	0.7306	0.0439	0.0141	30	6.00	121.00	0
122	Aspartic (%)	0859	0.9533	0.0262	1.165	0.0397	0.0146	29	-5.32	122.00	0
122	Aspartic (%)	0626	1.102	0.0230	1.165	0.0397	0.0146	29	-1.59	122.05	0
122	Aspartic (%)	0870	1.113	0.0193	1.165	0.0397	0.0146	29	-1.30	122.00	0
122	Aspartic (%)	0504	1.120	0.0000	1.165	0.0397	0.0146	29	-1.12	122.00	0
122	Aspartic (%)	0407	1.127	0.0540	1.165	0.0397	0.0146	29	-0.95	122.05	0
122	Aspartic (%)	0682	1.143	0.0000	1.165	0.0397	0.0146	29	-0.54	122.00	0
122	Aspartic (%)	0872	1.145	0.0000	1.165	0.0397	0.0146	29	-0.49	122.00	0
122	Aspartic (%)	0968	1.148	0.0100	1.165	0.0397	0.0146	29	-0.42	122.00	0
122	Aspartic (%)	0644	1.149	0.0310	1.165	0.0397	0.0146	29	-0.41	122.00	0
122	Aspartic (%)	0675	1.150	0.0200	1.165	0.0397	0.0146	29	-0.37	122.00	0
122	Aspartic (%)	0148	1.150	0.0200	1.165	0.0397	0.0146	29	-0.37	122.05	0
122	Aspartic (%)	0610	1.150	0.0000	1.165	0.0397	0.0146	29	-0.37	122.05	0
122	Aspartic (%)	0227	1.155	0.0100	1.165	0.0397	0.0146	29	-0.24	122.00	0
122	Aspartic (%)	0910	1.155	0.0300	1.165	0.0397	0.0146	29	-0.24	122.00	0
122	Aspartic (%)	0676	1.155	0.0300	1.165	0.0397	0.0146	29	-0.24	122.05	0
122	Aspartic (%)	0354	1.156	0.0210	1.165	0.0397	0.0146	29	-0.23	122.00	0
122	Aspartic (%)	0878	1.165	0.0060	1.165	0.0397	0.0146	29	0.01	122.00	0
122	Aspartic (%)	2059	1.171	0.0010	1.165	0.0397	0.0146	29	0.15	122.00	0
122	Aspartic (%)	0868	1.173	0.0060	1.165	0.0397	0.0146	29	0.21	122.00	0
122	Aspartic (%)	0571	1.178	0.0090	1.165	0.0397	0.0146	29	0.33	122.00	0
122	Aspartic (%)	0098	1.182	0.0040	1.165	0.0397	0.0146	29	0.44	122.02	0
122	Aspartic (%)	0918	1.188	0.0157	1.165	0.0397	0.0146	29	0.59	122.00	0
122	Aspartic (%)	2196	1.196	0.0000	1.165	0.0397	0.0146	29	0.79	122.05	0
122	Aspartic (%)	0171	1.200	0.0110	1.165	0.0397	0.0146	29	0.88	122.00	0
122	Aspartic (%)	0353	1.200	0.0000	1.165	0.0397	0.0146	29	0.89	122.00	0
122	Aspartic (%)	0226	1.234	0.0159	1.165	0.0397	0.0146	29	1.75	122.00	0
122	Aspartic (%)	0619	1.245	0.0100	1.165	0.0397	0.0146	29	2.03	122.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
122	Aspartic (%)	2188	1.265	0.0500	1.165	0.0397	0.0146	29	2.53	122.05	0
122	Aspartic (%)	2195	1.540	0.0000	1.165	0.0397	0.0146	29	9.45	122.05	0
122	Aspartic (%)	0853	1.605	0.2300	1.165	0.0397	0.0146	29	11.09	122.05	1
124	Cysteine/Cystine (%)	0853	0.0250	0.0100	0.2079	0.0262	0.0057	30	-6.97	124.05	0
124	Cysteine/Cystine (%)	0354	0.1620	0.0040	0.2079	0.0262	0.0057	30	-1.75	124.00	0
124	Cysteine/Cystine (%)	0870	0.1797	0.0001	0.2079	0.0262	0.0057	30	-1.08	124.00	0
124	Cysteine/Cystine (%)	2195	0.1800	0.0000	0.2079	0.0262	0.0057	30	-1.06	124.05	0
124	Cysteine/Cystine (%)	2196	0.1800	0.0000	0.2079	0.0262	0.0057	30	-1.06	124.05	0
124	Cysteine/Cystine (%)	0676	0.1850	0.0100	0.2079	0.0262	0.0057	30	-0.87	124.05	0
124	Cysteine/Cystine (%)	0859	0.1900	0.0044	0.2079	0.0262	0.0057	30	-0.68	124.00	0
124	Cysteine/Cystine (%)	0910	0.1900	0.0200	0.2079	0.0262	0.0057	30	-0.68	124.00	0
124	Cysteine/Cystine (%)	0227	0.1950	0.0100	0.2079	0.0262	0.0057	30	-0.49	124.00	0
124	Cysteine/Cystine (%)	0644	0.1950	0.0020	0.2079	0.0262	0.0057	30	-0.49	124.00	0
124	Cysteine/Cystine (%)	0504	0.2000	0.0200	0.2079	0.0262	0.0057	30	-0.30	124.00	0
124	Cysteine/Cystine (%)	0918	0.2000	0.0000	0.2079	0.0262	0.0057	30	-0.30	124.00	0
124	Cysteine/Cystine (%)	0723	0.2000	0.0000	0.2079	0.0262	0.0057	30	-0.30	124.05	0
124	Cysteine/Cystine (%)	0353	0.2050	0.0100	0.2079	0.0262	0.0057	30	-0.11	124.00	0
124	Cysteine/Cystine (%)	0878	0.2065	0.0030	0.2079	0.0262	0.0057	30	-0.05	124.00	0
124	Cysteine/Cystine (%)	0171	0.2070	0.0140	0.2079	0.0262	0.0057	30	-0.03	124.00	0
124	Cysteine/Cystine (%)	0571	0.2085	0.0030	0.2079	0.0262	0.0057	30	0.02	124.00	0
124	Cysteine/Cystine (%)	0848	0.2100	0.0000	0.2079	0.0262	0.0057	30	0.08	124.00	0
124	Cysteine/Cystine (%)	2059	0.2125	0.0070	0.2079	0.0262	0.0057	30	0.18	124.00	0
124	Cysteine/Cystine (%)	0098	0.2155	0.0050	0.2079	0.0262	0.0057	30	0.29	124.02	0
124	Cysteine/Cystine (%)	0968	0.2175	0.0010	0.2079	0.0262	0.0057	30	0.37	124.00	0
124	Cysteine/Cystine (%)	0872	0.2185	0.0030	0.2079	0.0262	0.0057	30	0.40	124.00	0
124	Cysteine/Cystine (%)	0868	0.2195	0.0130	0.2079	0.0262	0.0057	30	0.44	124.00	0
124	Cysteine/Cystine (%)	0610	0.2300	0.0000	0.2079	0.0262	0.0057	30	0.84	124.05	0
124	Cysteine/Cystine (%)	2188	0.2300	0.0000	0.2079	0.0262	0.0057	30	0.84	124.05	0
124	Cysteine/Cystine (%)	0226	0.2359	0.0003	0.2079	0.0262	0.0057	30	1.06	124.00	0
124	Cysteine/Cystine (%)	0675	0.2500	0.0200	0.2079	0.0262	0.0057	30	1.60	124.00	0
124	Cysteine/Cystine (%)	0682	0.2780	0.0000	0.2079	0.0262	0.0057	30	2.67	124.00	0
124	Cysteine/Cystine (%)	0619	0.2925	0.0110	0.2079	0.0262	0.0057	30	3.22	124.00	0
124	Cysteine/Cystine (%)	0407	0.3265	0.0010	0.2079	0.0262	0.0057	30	4.52	124.05	0
124	Cysteine/Cystine (%)	0148	0.1200	0.0400	0.2079	0.0262	0.0057	30	-3.35	124.05	1
125	Glutamic (%)	0859	1.646	0.0363	2.026	0.0873	0.0178	29	-4.35	125.00	0
125	Glutamic (%)	0171	1.737	0.0590	2.026	0.0873	0.0178	29	-3.31	125.00	0
125	Glutamic (%)	0626	1.898	0.0160	2.026	0.0873	0.0178	29	-1.46	125.05	0
125	Glutamic (%)	0407	1.932	0.0120	2.026	0.0873	0.0178	29	-1.07	125.05	0
125	Glutamic (%)	0870	1.961	0.0068	2.026	0.0873	0.0178	29	-0.74	125.00	0
125	Glutamic (%)	0675	1.980	0.0200	2.026	0.0873	0.0178	29	-0.52	125.00	0
125	Glutamic (%)	0354	1.988	0.0060	2.026	0.0873	0.0178	29	-0.43	125.00	0
125	Glutamic (%)	0676	1.990	0.0600	2.026	0.0873	0.0178	29	-0.41	125.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
125	Glutamic (%)	0227	1.995	0.0100	2.026	0.0873	0.0178	29	-0.35	125.00	0
125	Glutamic (%)	0504	1.995	0.0100	2.026	0.0873	0.0178	29	-0.35	125.00	0
125	Glutamic (%)	0644	2.002	0.0310	2.026	0.0873	0.0178	29	-0.28	125.00	0
125	Glutamic (%)	0968	2.005	0.0240	2.026	0.0873	0.0178	29	-0.24	125.00	0
125	Glutamic (%)	2059	2.006	0.0010	2.026	0.0873	0.0178	29	-0.23	125.00	0
125	Glutamic (%)	2196	2.006	0.0000	2.026	0.0873	0.0178	29	-0.22	125.05	0
125	Glutamic (%)	0571	2.011	0.0060	2.026	0.0873	0.0178	29	-0.17	125.00	0
125	Glutamic (%)	0872	2.019	0.0050	2.026	0.0873	0.0178	29	-0.08	125.00	0
125	Glutamic (%)	2195	2.020	0.0000	2.026	0.0873	0.0178	29	-0.06	125.05	0
125	Glutamic (%)	0910	2.025	0.0300	2.026	0.0873	0.0178	29	-0.01	125.00	0
125	Glutamic (%)	0610	2.025	0.0100	2.026	0.0873	0.0178	29	-0.01	125.05	0
125	Glutamic (%)	0878	2.035	0.0150	2.026	0.0873	0.0178	29	0.10	125.00	0
125	Glutamic (%)	0682	2.051	0.0000	2.026	0.0873	0.0178	29	0.29	125.00	0
125	Glutamic (%)	0918	2.062	0.0354	2.026	0.0873	0.0178	29	0.42	125.00	0
125	Glutamic (%)	0098	2.063	0.0060	2.026	0.0873	0.0178	29	0.43	125.02	0
125	Glutamic (%)	0868	2.124	0.0120	2.026	0.0873	0.0178	29	1.13	125.00	0
125	Glutamic (%)	0148	2.135	0.0100	2.026	0.0873	0.0178	29	1.25	125.05	0
125	Glutamic (%)	2188	2.160	0.0400	2.026	0.0873	0.0178	29	1.54	125.05	0
125	Glutamic (%)	0353	2.210	0.0000	2.026	0.0873	0.0178	29	2.11	125.00	0
125	Glutamic (%)	0226	2.211	0.0147	2.026	0.0873	0.0178	29	2.12	125.00	0
125	Glutamic (%)	0619	2.260	0.0400	2.026	0.0873	0.0178	29	2.69	125.00	0
125	Glutamic (%)	0853	2.910	0.3600	2.026	0.0873	0.0178	29	10.13	125.05	1
126	Glycine (%)	0859	0.5311	0.0117	0.6459	0.0302	0.0100	30	-3.80	126.00	0
126	Glycine (%)	0870	0.5889	0.0188	0.6459	0.0302	0.0100	30	-1.89	126.00	0
126	Glycine (%)	0910	0.5900	0.0400	0.6459	0.0302	0.0100	30	-1.85	126.00	0
126	Glycine (%)	0354	0.6005	0.0250	0.6459	0.0302	0.0100	30	-1.50	126.00	0
126	Glycine (%)	0171	0.6105	0.0030	0.6459	0.0302	0.0100	30	-1.17	126.00	0
126	Glycine (%)	0675	0.6200	0.0200	0.6459	0.0302	0.0100	30	-0.86	126.00	0
126	Glycine (%)	0968	0.6255	0.0050	0.6459	0.0302	0.0100	30	-0.68	126.00	0
126	Glycine (%)	0868	0.6260	0.0020	0.6459	0.0302	0.0100	30	-0.66	126.00	0
126	Glycine (%)	0148	0.6350	0.0100	0.6459	0.0302	0.0100	30	-0.36	126.05	0
126	Glycine (%)	0644	0.6370	0.0040	0.6459	0.0302	0.0100	30	-0.30	126.00	0
126	Glycine (%)	0676	0.6400	0.0000	0.6459	0.0302	0.0100	30	-0.20	126.05	0
126	Glycine (%)	0098	0.6420	0.0020	0.6459	0.0302	0.0100	30	-0.13	126.02	0
126	Glycine (%)	0227	0.6450	0.0100	0.6459	0.0302	0.0100	30	-0.03	126.00	0
126	Glycine (%)	0504	0.6450	0.0100	0.6459	0.0302	0.0100	30	-0.03	126.00	0
126	Glycine (%)	0571	0.6485	0.0010	0.6459	0.0302	0.0100	30	0.08	126.00	0
126	Glycine (%)	0872	0.6485	0.0110	0.6459	0.0302	0.0100	30	0.08	126.00	0
126	Glycine (%)	0878	0.6495	0.0110	0.6459	0.0302	0.0100	30	0.12	126.00	0
126	Glycine (%)	0682	0.6500	0.0000	0.6459	0.0302	0.0100	30	0.13	126.00	0
126	Glycine (%)	2059	0.6500	0.0000	0.6459	0.0302	0.0100	30	0.13	126.00	0
126	Glycine (%)	2196	0.6580	0.0000	0.6459	0.0302	0.0100	30	0.40	126.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
126	Glycine (%)	0610	0.6600	0.0000	0.6459	0.0302	0.0100	30	0.47	126.05	0
126	Glycine (%)	0918	0.6623	0.0060	0.6459	0.0302	0.0100	30	0.54	126.00	0
126	Glycine (%)	0353	0.6700	0.0000	0.6459	0.0302	0.0100	30	0.80	126.00	0
126	Glycine (%)	0407	0.6700	0.0060	0.6459	0.0302	0.0100	30	0.80	126.05	0
126	Glycine (%)	0619	0.6710	0.0160	0.6459	0.0302	0.0100	30	0.83	126.00	0
126	Glycine (%)	0626	0.6745	0.0010	0.6459	0.0302	0.0100	30	0.95	126.05	0
126	Glycine (%)	0226	0.6749	0.0070	0.6459	0.0302	0.0100	30	0.96	126.00	0
126	Glycine (%)	0853	0.6800	0.0400	0.6459	0.0302	0.0100	30	1.13	126.05	0
126	Glycine (%)	2188	0.7200	0.0400	0.6459	0.0302	0.0100	30	2.45	126.05	0
126	Glycine (%)	2195	0.7300	0.0000	0.6459	0.0302	0.0100	30	2.78	126.05	0
127	Histidine (%)	0853	0.1200	0.0200	0.2827	0.0176	0.0056	29	-9.27	127.05	0
127	Histidine (%)	0870	0.1839	0.0053	0.2827	0.0176	0.0056	29	-5.63	127.00	0
127	Histidine (%)	0859	0.2222	0.0060	0.2827	0.0176	0.0056	29	-3.45	127.00	0
127	Histidine (%)	0354	0.2425	0.0010	0.2827	0.0176	0.0056	29	-2.29	127.00	0
127	Histidine (%)	0610	0.2650	0.0100	0.2827	0.0176	0.0056	29	-1.01	127.05	0
127	Histidine (%)	2196	0.2710	0.0000	0.2827	0.0176	0.0056	29	-0.67	127.05	0
127	Histidine (%)	0226	0.2715	0.0072	0.2827	0.0176	0.0056	29	-0.64	127.00	0
127	Histidine (%)	2188	0.2750	0.0100	0.2827	0.0176	0.0056	29	-0.44	127.05	0
127	Histidine (%)	0968	0.2760	0.0020	0.2827	0.0176	0.0056	29	-0.38	127.00	0
127	Histidine (%)	0504	0.2800	0.0000	0.2827	0.0176	0.0056	29	-0.16	127.00	0
127	Histidine (%)	0910	0.2800	0.0200	0.2827	0.0176	0.0056	29	-0.16	127.00	0
127	Histidine (%)	0148	0.2800	0.0000	0.2827	0.0176	0.0056	29	-0.16	127.05	0
127	Histidine (%)	0676	0.2800	0.0000	0.2827	0.0176	0.0056	29	-0.16	127.05	0
127	Histidine (%)	0878	0.2815	0.0010	0.2827	0.0176	0.0056	29	-0.07	127.00	0
127	Histidine (%)	0626	0.2825	0.0010	0.2827	0.0176	0.0056	29	-0.01	127.05	0
127	Histidine (%)	0098	0.2830	0.0040	0.2827	0.0176	0.0056	29	0.01	127.02	0
127	Histidine (%)	0571	0.2860	0.0040	0.2827	0.0176	0.0056	29	0.19	127.00	0
127	Histidine (%)	2059	0.2865	0.0010	0.2827	0.0176	0.0056	29	0.21	127.00	0
127	Histidine (%)	0918	0.2883	0.0040	0.2827	0.0176	0.0056	29	0.32	127.00	0
127	Histidine (%)	0872	0.2895	0.0030	0.2827	0.0176	0.0056	29	0.38	127.00	0
127	Histidine (%)	0353	0.2900	0.0000	0.2827	0.0176	0.0056	29	0.41	127.00	0
127	Histidine (%)	0682	0.2910	0.0000	0.2827	0.0176	0.0056	29	0.47	127.00	0
127	Histidine (%)	0644	0.2915	0.0010	0.2827	0.0176	0.0056	29	0.50	127.00	0
127	Histidine (%)	0868	0.2985	0.0030	0.2827	0.0176	0.0056	29	0.90	127.00	0
127	Histidine (%)	0227	0.3000	0.0000	0.2827	0.0176	0.0056	29	0.98	127.00	0
127	Histidine (%)	2195	0.3000	0.0000	0.2827	0.0176	0.0056	29	0.98	127.05	0
127	Histidine (%)	0619	0.3610	0.0080	0.2827	0.0176	0.0056	29	4.46	127.00	0
127	Histidine (%)	0675	0.3700	0.0200	0.2827	0.0176	0.0056	29	4.97	127.00	0
127	Histidine (%)	0407	0.3710	0.0320	0.2827	0.0176	0.0056	29	5.02	127.05	0
127	Histidine (%)	0171	0.6465	0.0350	0.2827	0.0176	0.0056	29	20.71	127.00	1
128	Isoleucine (%)	0870	0.2644	0.0088	0.4339	0.0370	0.0098	29	-4.58	128.00	0
128	Isoleucine (%)	0853	0.3300	0.0400	0.4339	0.0370	0.0098	29	-2.81	128.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
128	Isoleucine (%)	0353	0.3700	0.0000	0.4339	0.0370	0.0098	29	-1.73	128.00	0
128	Isoleucine (%)	0407	0.3800	0.0100	0.4339	0.0370	0.0098	29	-1.46	128.05	0
128	Isoleucine (%)	0626	0.3865	0.0010	0.4339	0.0370	0.0098	29	-1.28	128.05	0
128	Isoleucine (%)	0675	0.4000	0.0200	0.4339	0.0370	0.0098	29	-0.92	128.00	0
128	Isoleucine (%)	0171	0.4140	0.0160	0.4339	0.0370	0.0098	29	-0.54	128.00	0
128	Isoleucine (%)	0354	0.4150	0.0040	0.4339	0.0370	0.0098	29	-0.51	128.00	0
128	Isoleucine (%)	0868	0.4245	0.0310	0.4339	0.0370	0.0098	29	-0.25	128.00	0
128	Isoleucine (%)	0968	0.4245	0.0070	0.4339	0.0370	0.0098	29	-0.25	128.00	0
128	Isoleucine (%)	0619	0.4275	0.0110	0.4339	0.0370	0.0098	29	-0.17	128.00	0
128	Isoleucine (%)	0682	0.4290	0.0000	0.4339	0.0370	0.0098	29	-0.13	128.00	0
128	Isoleucine (%)	0676	0.4300	0.0200	0.4339	0.0370	0.0098	29	-0.11	128.05	0
128	Isoleucine (%)	0859	0.4371	0.0085	0.4339	0.0370	0.0098	29	0.08	128.00	0
128	Isoleucine (%)	0918	0.4377	0.0082	0.4339	0.0370	0.0098	29	0.10	128.00	0
128	Isoleucine (%)	2196	0.4420	0.0000	0.4339	0.0370	0.0098	29	0.22	128.05	0
128	Isoleucine (%)	0878	0.4445	0.0050	0.4339	0.0370	0.0098	29	0.29	128.00	0
128	Isoleucine (%)	0872	0.4465	0.0170	0.4339	0.0370	0.0098	29	0.34	128.00	0
128	Isoleucine (%)	0098	0.4470	0.0140	0.4339	0.0370	0.0098	29	0.35	128.02	0
128	Isoleucine (%)	0644	0.4490	0.0160	0.4339	0.0370	0.0098	29	0.41	128.00	0
128	Isoleucine (%)	0610	0.4500	0.0000	0.4339	0.0370	0.0098	29	0.44	128.05	0
128	Isoleucine (%)	0227	0.4550	0.0100	0.4339	0.0370	0.0098	29	0.57	128.00	0
128	Isoleucine (%)	0910	0.4550	0.0100	0.4339	0.0370	0.0098	29	0.57	128.00	0
128	Isoleucine (%)	0571	0.4570	0.0060	0.4339	0.0370	0.0098	29	0.62	128.00	0
128	Isoleucine (%)	2059	0.4625	0.0010	0.4339	0.0370	0.0098	29	0.77	128.00	0
128	Isoleucine (%)	0504	0.4750	0.0100	0.4339	0.0370	0.0098	29	1.11	128.00	0
128	Isoleucine (%)	2195	0.4800	0.0000	0.4339	0.0370	0.0098	29	1.25	128.05	0
128	Isoleucine (%)	0148	0.4900	0.0000	0.4339	0.0370	0.0098	29	1.52	128.05	0
128	Isoleucine (%)	2188	0.5050	0.0100	0.4339	0.0370	0.0098	29	1.92	128.05	0
128	Isoleucine (%)	0226	0.4476	0.0511	0.4339	0.0370	0.0098	29	0.37	128.00	1
129	Leucine (%)	0870	0.6834	0.0196	0.8200	0.0384	0.0105	29	-3.56	129.00	0
129	Leucine (%)	0859	0.7649	0.0220	0.8200	0.0384	0.0105	29	-1.43	129.00	0
129	Leucine (%)	0626	0.7705	0.0090	0.8200	0.0384	0.0105	29	-1.29	129.05	0
129	Leucine (%)	2196	0.7710	0.0000	0.8200	0.0384	0.0105	29	-1.28	129.05	0
129	Leucine (%)	0968	0.7735	0.0010	0.8200	0.0384	0.0105	29	-1.21	129.00	0
129	Leucine (%)	0675	0.7800	0.0200	0.8200	0.0384	0.0105	29	-1.04	129.00	0
129	Leucine (%)	0354	0.7820	0.0200	0.8200	0.0384	0.0105	29	-0.99	129.00	0
129	Leucine (%)	0682	0.7910	0.0000	0.8200	0.0384	0.0105	29	-0.75	129.00	0
129	Leucine (%)	2195	0.8000	0.0000	0.8200	0.0384	0.0105	29	-0.52	129.05	0
129	Leucine (%)	0171	0.8075	0.0210	0.8200	0.0384	0.0105	29	-0.32	129.00	0
129	Leucine (%)	0407	0.8110	0.0060	0.8200	0.0384	0.0105	29	-0.23	129.05	0
129	Leucine (%)	0910	0.8200	0.0000	0.8200	0.0384	0.0105	29	0.00	129.00	0
129	Leucine (%)	0676	0.8250	0.0300	0.8200	0.0384	0.0105	29	0.13	129.05	0
129	Leucine (%)	0878	0.8280	0.0100	0.8200	0.0384	0.0105	29	0.21	129.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
129	Leucine (%)	0868	0.8300	0.0140	0.8200	0.0384	0.0105	29	0.26	129.00	0
129	Leucine (%)	0148	0.8300	0.0000	0.8200	0.0384	0.0105	29	0.26	129.05	0
129	Leucine (%)	0644	0.8315	0.0170	0.8200	0.0384	0.0105	29	0.30	129.00	0
129	Leucine (%)	0872	0.8330	0.0140	0.8200	0.0384	0.0105	29	0.34	129.00	0
129	Leucine (%)	0098	0.8345	0.0050	0.8200	0.0384	0.0105	29	0.38	129.02	0
129	Leucine (%)	0227	0.8350	0.0100	0.8200	0.0384	0.0105	29	0.39	129.00	0
129	Leucine (%)	0610	0.8350	0.0100	0.8200	0.0384	0.0105	29	0.39	129.05	0
129	Leucine (%)	2059	0.8395	0.0030	0.8200	0.0384	0.0105	29	0.51	129.00	0
129	Leucine (%)	0571	0.8430	0.0060	0.8200	0.0384	0.0105	29	0.60	129.00	0
129	Leucine (%)	0504	0.8500	0.0000	0.8200	0.0384	0.0105	29	0.78	129.00	0
129	Leucine (%)	0353	0.8550	0.0100	0.8200	0.0384	0.0105	29	0.91	129.00	0
129	Leucine (%)	0918	0.8565	0.0031	0.8200	0.0384	0.0105	29	0.95	129.00	0
129	Leucine (%)	0226	0.8644	0.0302	0.8200	0.0384	0.0105	29	1.16	129.00	0
129	Leucine (%)	0619	0.8870	0.0140	0.8200	0.0384	0.0105	29	1.75	129.00	0
129	Leucine (%)	2188	0.9250	0.0100	0.8200	0.0384	0.0105	29	2.74	129.05	0
129	Leucine (%)	0853	0.8850	0.0900	0.8200	0.0384	0.0105	29	1.69	129.05	1
130	L-Lysine (%)	0171	0.5515	0.0210	0.6970	0.0328	0.0108	34	-4.44	130.00	0
130	L-Lysine (%)	0859	0.5646	0.0086	0.6970	0.0328	0.0108	34	-4.04	130.00	0
130	L-Lysine (%)	2196	0.6570	0.0000	0.6970	0.0328	0.0108	34	-1.22	130.05	0
130	L-Lysine (%)	0870	0.6588	0.0042	0.6970	0.0328	0.0108	34	-1.17	130.00	0
130	L-Lysine (%)	0407	0.6615	0.0190	0.6970	0.0328	0.0108	34	-1.08	130.05	0
130	L-Lysine (%)	0910	0.6650	0.0100	0.6970	0.0328	0.0108	34	-0.98	130.00	0
130	L-Lysine (%)	0354	0.6730	0.0120	0.6970	0.0328	0.0108	34	-0.73	130.00	0
130	L-Lysine (%)	0626	0.6765	0.0030	0.6970	0.0328	0.0108	34	-0.63	130.05	0
130	L-Lysine (%)	0939	0.6800	0.0200	0.6970	0.0328	0.0108	34	-0.52	130.99	0
130	L-Lysine (%)	0918	0.6806	0.0067	0.6970	0.0328	0.0108	34	-0.50	130.00	0
130	L-Lysine (%)	0878	0.6825	0.0070	0.6970	0.0328	0.0108	34	-0.44	130.00	0
130	L-Lysine (%)	0610	0.6850	0.0100	0.6970	0.0328	0.0108	34	-0.37	130.05	0
130	L-Lysine (%)	0968	0.6860	0.0100	0.6970	0.0328	0.0108	34	-0.34	130.00	0
130	L-Lysine (%)	0872	0.6865	0.0050	0.6970	0.0328	0.0108	34	-0.32	130.00	0
130	L-Lysine (%)	0644	0.6870	0.0040	0.6970	0.0328	0.0108	34	-0.30	130.00	0
130	L-Lysine (%)	0148	0.6900	0.0200	0.6970	0.0328	0.0108	34	-0.21	130.05	0
130	L-Lysine (%)	0226	0.6928	0.0352	0.6970	0.0328	0.0108	34	-0.13	130.00	0
130	L-Lysine (%)	0353	0.7000	0.0000	0.6970	0.0328	0.0108	34	0.09	130.00	0
130	L-Lysine (%)	0675	0.7000	0.0200	0.6970	0.0328	0.0108	34	0.09	130.00	0
130	L-Lysine (%)	2195	0.7000	0.0000	0.6970	0.0328	0.0108	34	0.09	130.05	0
130	L-Lysine (%)	0014	0.7005	0.0530	0.6970	0.0328	0.0108	34	0.11	130.05	0
130	L-Lysine (%)	0571	0.7010	0.0020	0.6970	0.0328	0.0108	34	0.12	130.00	0
130	L-Lysine (%)	0868	0.7030	0.0140	0.6970	0.0328	0.0108	34	0.18	130.00	0
130	L-Lysine (%)	2059	0.7070	0.0040	0.6970	0.0328	0.0108	34	0.31	130.00	0
130	L-Lysine (%)	0098	0.7115	0.0030	0.6970	0.0328	0.0108	34	0.44	130.02	0
130	L-Lysine (%)	0676	0.7150	0.0100	0.6970	0.0328	0.0108	34	0.55	130.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
130	L-Lysine (%)	2053	0.7150	0.0100	0.6970	0.0328	0.0108	34	0.55	130.99	0
130	L-Lysine (%)	2146	0.7250	0.0100	0.6970	0.0328	0.0108	34	0.86	130.99	0
130	L-Lysine (%)	0682	0.7310	0.0000	0.6970	0.0328	0.0108	34	1.04	130.00	0
130	L-Lysine (%)	0504	0.7500	0.0000	0.6970	0.0328	0.0108	34	1.62	130.00	0
130	L-Lysine (%)	2145	0.7550	0.0100	0.6970	0.0328	0.0108	34	1.77	130.01	0
130	L-Lysine (%)	2188	0.7700	0.0000	0.6970	0.0328	0.0108	34	2.23	130.05	0
130	L-Lysine (%)	0227	0.7750	0.0100	0.6970	0.0328	0.0108	34	2.38	130.00	0
130	L-Lysine (%)	0619	0.9470	0.0260	0.6970	0.0328	0.0108	34	7.63	130.00	0
130	L-Lysine (%)	0853	0.7750	0.1300	0.6970	0.0328	0.0108	34	2.38	130.05	1
131	Methionine (%)	0853	0.0500	0.0200	0.1581	0.0215	0.0060	36	-5.02	131.05	0
131	Methionine (%)	0354	0.0805	0.0030	0.1581	0.0215	0.0060	36	-3.61	131.00	0
131	Methionine (%)	0619	0.0915	0.0010	0.1581	0.0215	0.0060	36	-3.09	131.00	0
131	Methionine (%)	0675	0.1300	0.0200	0.1581	0.0215	0.0060	36	-1.31	131.00	0
131	Methionine (%)	0848	0.1300	0.0000	0.1581	0.0215	0.0060	36	-1.31	131.00	0
131	Methionine (%)	0148	0.1300	0.0000	0.1581	0.0215	0.0060	36	-1.31	131.05	0
131	Methionine (%)	0859	0.1309	0.0032	0.1581	0.0215	0.0060	36	-1.26	131.00	0
131	Methionine (%)	2188	0.1350	0.0100	0.1581	0.0215	0.0060	36	-1.07	131.05	0
131	Methionine (%)	0626	0.1365	0.0070	0.1581	0.0215	0.0060	36	-1.00	131.05	0
131	Methionine (%)	0171	0.1495	0.0030	0.1581	0.0215	0.0060	36	-0.40	131.00	0
131	Methionine (%)	0353	0.1500	0.0000	0.1581	0.0215	0.0060	36	-0.38	131.00	0
131	Methionine (%)	0227	0.1550	0.0100	0.1581	0.0215	0.0060	36	-0.14	131.00	0
131	Methionine (%)	0504	0.1550	0.0100	0.1581	0.0215	0.0060	36	-0.14	131.00	0
131	Methionine (%)	0910	0.1550	0.0100	0.1581	0.0215	0.0060	36	-0.14	131.00	0
131	Methionine (%)	0870	0.1579	0.0054	0.1581	0.0215	0.0060	36	-0.01	131.00	0
131	Methionine (%)	0723	0.1600	0.0000	0.1581	0.0215	0.0060	36	0.09	131.05	0
131	Methionine (%)	0872	0.1605	0.0010	0.1581	0.0215	0.0060	36	0.11	131.00	0
131	Methionine (%)	0968	0.1605	0.0090	0.1581	0.0215	0.0060	36	0.11	131.00	0
131	Methionine (%)	0098	0.1610	0.0020	0.1581	0.0215	0.0060	36	0.13	131.02	0
131	Methionine (%)	2146	0.1650	0.0100	0.1581	0.0215	0.0060	36	0.32	131.99	0
131	Methionine (%)	0571	0.1660	0.0020	0.1581	0.0215	0.0060	36	0.37	131.00	0
131	Methionine (%)	0868	0.1665	0.0030	0.1581	0.0215	0.0060	36	0.39	131.00	0
131	Methionine (%)	2059	0.1690	0.0000	0.1581	0.0215	0.0060	36	0.51	131.00	0
131	Methionine (%)	0878	0.1695	0.0010	0.1581	0.0215	0.0060	36	0.53	131.00	0
131	Methionine (%)	0226	0.1696	0.0070	0.1581	0.0215	0.0060	36	0.53	131.00	0
131	Methionine (%)	0918	0.1700	0.0000	0.1581	0.0215	0.0060	36	0.55	131.00	0
131	Methionine (%)	0610	0.1700	0.0000	0.1581	0.0215	0.0060	36	0.55	131.05	0
131	Methionine (%)	0014	0.1710	0.0340	0.1581	0.0215	0.0060	36	0.60	131.05	0
131	Methionine (%)	0407	0.1729	0.0002	0.1581	0.0215	0.0060	36	0.69	131.05	0
131	Methionine (%)	0644	0.1740	0.0040	0.1581	0.0215	0.0060	36	0.74	131.00	0
131	Methionine (%)	0682	0.1740	0.0000	0.1581	0.0215	0.0060	36	0.74	131.00	0
131	Methionine (%)	2196	0.1800	0.0000	0.1581	0.0215	0.0060	36	1.02	131.05	0
131	Methionine (%)	2145	0.1800	0.0000	0.1581	0.0215	0.0060	36	1.02	131.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
131	Methionine (%)	0676	0.1850	0.0100	0.1581	0.0215	0.0060	36	1.25	131.05	0
131	Methionine (%)	0939	0.1850	0.0300	0.1581	0.0215	0.0060	36	1.25	131.99	0
131	Methionine (%)	2195	0.2200	0.0000	0.1581	0.0215	0.0060	36	2.87	131.05	0
131	Methionine (%)	2053	0.1800	0.0400	0.1581	0.0215	0.0060	36	1.02	131.99	1
132	Phenylalanine (%)	0870	0.4526	0.0124	0.5331	0.0265	0.0099	30	-3.03	132.00	0
132	Phenylalanine (%)	0853	0.4700	0.0200	0.5331	0.0265	0.0099	30	-2.38	132.05	0
132	Phenylalanine (%)	2196	0.4980	0.0000	0.5331	0.0265	0.0099	30	-1.32	132.05	0
132	Phenylalanine (%)	0878	0.4985	0.0110	0.5331	0.0265	0.0099	30	-1.31	132.00	0
132	Phenylalanine (%)	0354	0.5030	0.0300	0.5331	0.0265	0.0099	30	-1.14	132.00	0
132	Phenylalanine (%)	0682	0.5100	0.0000	0.5331	0.0265	0.0099	30	-0.87	132.00	0
132	Phenylalanine (%)	0675	0.5200	0.0200	0.5331	0.0265	0.0099	30	-0.50	132.00	0
132	Phenylalanine (%)	2195	0.5200	0.0000	0.5331	0.0265	0.0099	30	-0.50	132.05	0
132	Phenylalanine (%)	0148	0.5250	0.0100	0.5331	0.0265	0.0099	30	-0.31	132.05	0
132	Phenylalanine (%)	0610	0.5250	0.0100	0.5331	0.0265	0.0099	30	-0.31	132.05	0
132	Phenylalanine (%)	0676	0.5250	0.0100	0.5331	0.0265	0.0099	30	-0.31	132.05	0
132	Phenylalanine (%)	0868	0.5260	0.0100	0.5331	0.0265	0.0099	30	-0.27	132.00	0
132	Phenylalanine (%)	0407	0.5265	0.0150	0.5331	0.0265	0.0099	30	-0.25	132.05	0
132	Phenylalanine (%)	0872	0.5270	0.0080	0.5331	0.0265	0.0099	30	-0.23	132.00	0
132	Phenylalanine (%)	0859	0.5277	0.0087	0.5331	0.0265	0.0099	30	-0.21	132.00	0
132	Phenylalanine (%)	0171	0.5285	0.0030	0.5331	0.0265	0.0099	30	-0.17	132.00	0
132	Phenylalanine (%)	0227	0.5350	0.0100	0.5331	0.0265	0.0099	30	0.07	132.00	0
132	Phenylalanine (%)	0098	0.5370	0.0040	0.5331	0.0265	0.0099	30	0.15	132.02	0
132	Phenylalanine (%)	0968	0.5395	0.0110	0.5331	0.0265	0.0099	30	0.24	132.00	0
132	Phenylalanine (%)	0626	0.5395	0.0030	0.5331	0.0265	0.0099	30	0.24	132.05	0
132	Phenylalanine (%)	0918	0.5441	0.0078	0.5331	0.0265	0.0099	30	0.41	132.00	0
132	Phenylalanine (%)	0226	0.5470	0.0314	0.5331	0.0265	0.0099	30	0.52	132.00	0
132	Phenylalanine (%)	0910	0.5500	0.0000	0.5331	0.0265	0.0099	30	0.63	132.00	0
132	Phenylalanine (%)	2059	0.5565	0.0030	0.5331	0.0265	0.0099	30	0.88	132.00	0
132	Phenylalanine (%)	0644	0.5585	0.0150	0.5331	0.0265	0.0099	30	0.96	132.00	0
132	Phenylalanine (%)	2188	0.5600	0.0200	0.5331	0.0265	0.0099	30	1.01	132.05	0
132	Phenylalanine (%)	0504	0.5650	0.0100	0.5331	0.0265	0.0099	30	1.20	132.00	0
132	Phenylalanine (%)	0571	0.5695	0.0030	0.5331	0.0265	0.0099	30	1.37	132.00	0
132	Phenylalanine (%)	0353	0.6300	0.0000	0.5331	0.0265	0.0099	30	3.65	132.00	0
132	Phenylalanine (%)	0619	0.7330	0.0100	0.5331	0.0265	0.0099	30	7.53	132.00	0
133	Proline (%)	0859	0.6714	0.0110	0.8327	0.0636	0.0171	28	-2.53	133.00	0
133	Proline (%)	0870	0.7002	0.0185	0.8327	0.0636	0.0171	28	-2.08	133.00	0
133	Proline (%)	0354	0.7390	0.0120	0.8327	0.0636	0.0171	28	-1.47	133.00	0
133	Proline (%)	0910	0.7400	0.0000	0.8327	0.0636	0.0171	28	-1.46	133.00	0
133	Proline (%)	0878	0.7885	0.0130	0.8327	0.0636	0.0171	28	-0.70	133.00	0
133	Proline (%)	2196	0.7970	0.0000	0.8327	0.0636	0.0171	28	-0.56	133.05	0
133	Proline (%)	0407	0.8000	0.0040	0.8327	0.0636	0.0171	28	-0.51	133.05	0
133	Proline (%)	0868	0.8075	0.0050	0.8327	0.0636	0.0171	28	-0.40	133.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
133	Proline (%)	0675	0.8100	0.0200	0.8327	0.0636	0.0171	28	-0.36	133.00	0
133	Proline (%)	0644	0.8115	0.0230	0.8327	0.0636	0.0171	28	-0.33	133.00	0
133	Proline (%)	0171	0.8135	0.0410	0.8327	0.0636	0.0171	28	-0.30	133.00	0
133	Proline (%)	0227	0.8150	0.0300	0.8327	0.0636	0.0171	28	-0.28	133.00	0
133	Proline (%)	0148	0.8250	0.0100	0.8327	0.0636	0.0171	28	-0.12	133.05	0
133	Proline (%)	0610	0.8300	0.0000	0.8327	0.0636	0.0171	28	-0.04	133.05	0
133	Proline (%)	0571	0.8310	0.0260	0.8327	0.0636	0.0171	28	-0.03	133.00	0
133	Proline (%)	2059	0.8335	0.0030	0.8327	0.0636	0.0171	28	0.01	133.00	0
133	Proline (%)	0504	0.8400	0.0000	0.8327	0.0636	0.0171	28	0.11	133.00	0
133	Proline (%)	0626	0.8455	0.0070	0.8327	0.0636	0.0171	28	0.20	133.05	0
133	Proline (%)	0682	0.8530	0.0000	0.8327	0.0636	0.0171	28	0.32	133.00	0
133	Proline (%)	0918	0.8610	0.0048	0.8327	0.0636	0.0171	28	0.44	133.00	0
133	Proline (%)	2195	0.8700	0.0000	0.8327	0.0636	0.0171	28	0.59	133.05	0
133	Proline (%)	0872	0.8705	0.0350	0.8327	0.0636	0.0171	28	0.59	133.00	0
133	Proline (%)	0676	0.8750	0.0100	0.8327	0.0636	0.0171	28	0.66	133.05	0
133	Proline (%)	0353	0.8900	0.0000	0.8327	0.0636	0.0171	28	0.90	133.00	0
133	Proline (%)	0968	0.9110	0.0460	0.8327	0.0636	0.0171	28	1.23	133.00	0
133	Proline (%)	0853	0.9300	0.1000	0.8327	0.0636	0.0171	28	1.53	133.05	0
133	Proline (%)	2188	0.9800	0.0400	0.8327	0.0636	0.0171	28	2.31	133.05	0
133	Proline (%)	0619	1.330	0.0200	0.8327	0.0636	0.0171	28	7.81	133.00	0
133	Proline (%)	0226	1.148	0.1085	0.8327	0.0636	0.0171	28	4.95	133.00	1
134	Serine (%)	0859	0.4624	0.0074	0.5582	0.0295	0.0107	29	-3.25	134.00	0
134	Serine (%)	2195	0.4800	0.0000	0.5582	0.0295	0.0107	29	-2.65	134.05	0
134	Serine (%)	0504	0.5150	0.0300	0.5582	0.0295	0.0107	29	-1.47	134.00	0
134	Serine (%)	0626	0.5195	0.0030	0.5582	0.0295	0.0107	29	-1.31	134.05	0
134	Serine (%)	0098	0.5215	0.0150	0.5582	0.0295	0.0107	29	-1.24	134.02	0
134	Serine (%)	2196	0.5350	0.0000	0.5582	0.0295	0.0107	29	-0.79	134.05	0
134	Serine (%)	0644	0.5370	0.0120	0.5582	0.0295	0.0107	29	-0.72	134.00	0
134	Serine (%)	0682	0.5480	0.0000	0.5582	0.0295	0.0107	29	-0.35	134.00	0
134	Serine (%)	0610	0.5500	0.0000	0.5582	0.0295	0.0107	29	-0.28	134.05	0
134	Serine (%)	0878	0.5545	0.0030	0.5582	0.0295	0.0107	29	-0.13	134.00	0
134	Serine (%)	0227	0.5550	0.0100	0.5582	0.0295	0.0107	29	-0.11	134.00	0
134	Serine (%)	2059	0.5550	0.0040	0.5582	0.0295	0.0107	29	-0.11	134.00	0
134	Serine (%)	0968	0.5565	0.0010	0.5582	0.0295	0.0107	29	-0.06	134.00	0
134	Serine (%)	0918	0.5586	0.0180	0.5582	0.0295	0.0107	29	0.01	134.00	0
134	Serine (%)	0910	0.5600	0.0200	0.5582	0.0295	0.0107	29	0.06	134.00	0
134	Serine (%)	0870	0.5610	0.0125	0.5582	0.0295	0.0107	29	0.09	134.00	0
134	Serine (%)	0171	0.5615	0.0050	0.5582	0.0295	0.0107	29	0.11	134.00	0
134	Serine (%)	0354	0.5660	0.0020	0.5582	0.0295	0.0107	29	0.27	134.00	0
134	Serine (%)	0872	0.5660	0.0060	0.5582	0.0295	0.0107	29	0.27	134.00	0
134	Serine (%)	0571	0.5680	0.0060	0.5582	0.0295	0.0107	29	0.33	134.00	0
134	Serine (%)	0676	0.5700	0.0000	0.5582	0.0295	0.0107	29	0.40	134.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
134	Serine (%)	0675	0.5700	0.0200	0.5582	0.0295	0.0107	29	0.40	134.00	0
134	Serine (%)	0148	0.5700	0.0400	0.5582	0.0295	0.0107	29	0.40	134.05	0
134	Serine (%)	0853	0.5750	0.0500	0.5582	0.0295	0.0107	29	0.57	134.05	0
134	Serine (%)	0226	0.5768	0.0256	0.5582	0.0295	0.0107	29	0.63	134.00	0
134	Serine (%)	0868	0.6055	0.0090	0.5582	0.0295	0.0107	29	1.61	134.00	0
134	Serine (%)	0353	0.6300	0.0000	0.5582	0.0295	0.0107	29	2.44	134.00	0
134	Serine (%)	0619	0.6525	0.0010	0.5582	0.0295	0.0107	29	3.20	134.00	0
134	Serine (%)	2188	0.6550	0.0100	0.5582	0.0295	0.0107	29	3.28	134.05	0
134	Serine (%)	0407	0.5045	0.1270	0.5582	0.0295	0.0107	29	-1.82	134.05	1
135	Threonine (%)	0626	0.3590	0.0060	0.4567	0.0253	0.0075	29	-3.86	135.05	0
135	Threonine (%)	0859	0.3743	0.0097	0.4567	0.0253	0.0075	29	-3.26	135.00	0
135	Threonine (%)	2195	0.4000	0.0000	0.4567	0.0253	0.0075	29	-2.24	135.05	0
135	Threonine (%)	0870	0.4129	0.0040	0.4567	0.0253	0.0075	29	-1.73	135.00	0
135	Threonine (%)	0682	0.4250	0.0000	0.4567	0.0253	0.0075	29	-1.25	135.00	0
135	Threonine (%)	0407	0.4310	0.0200	0.4567	0.0253	0.0075	29	-1.01	135.05	0
135	Threonine (%)	0676	0.4450	0.0100	0.4567	0.0253	0.0075	29	-0.46	135.05	0
135	Threonine (%)	0148	0.4500	0.0000	0.4567	0.0253	0.0075	29	-0.26	135.05	0
135	Threonine (%)	0872	0.4515	0.0070	0.4567	0.0253	0.0075	29	-0.20	135.00	0
135	Threonine (%)	0878	0.4550	0.0180	0.4567	0.0253	0.0075	29	-0.07	135.00	0
135	Threonine (%)	0910	0.4550	0.0100	0.4567	0.0253	0.0075	29	-0.07	135.00	0
135	Threonine (%)	0644	0.4570	0.0180	0.4567	0.0253	0.0075	29	0.01	135.00	0
135	Threonine (%)	0868	0.4580	0.0000	0.4567	0.0253	0.0075	29	0.05	135.00	0
135	Threonine (%)	0968	0.4580	0.0160	0.4567	0.0253	0.0075	29	0.05	135.00	0
135	Threonine (%)	2059	0.4595	0.0030	0.4567	0.0253	0.0075	29	0.11	135.00	0
135	Threonine (%)	0675	0.4600	0.0200	0.4567	0.0253	0.0075	29	0.13	135.00	0
135	Threonine (%)	0227	0.4600	0.0000	0.4567	0.0253	0.0075	29	0.13	135.00	0
135	Threonine (%)	0610	0.4600	0.0000	0.4567	0.0253	0.0075	29	0.13	135.05	0
135	Threonine (%)	0354	0.4625	0.0030	0.4567	0.0253	0.0075	29	0.23	135.00	0
135	Threonine (%)	0226	0.4644	0.0099	0.4567	0.0253	0.0075	29	0.30	135.00	0
135	Threonine (%)	0504	0.4650	0.0100	0.4567	0.0253	0.0075	29	0.33	135.00	0
135	Threonine (%)	0571	0.4650	0.0000	0.4567	0.0253	0.0075	29	0.33	135.00	0
135	Threonine (%)	0098	0.4665	0.0090	0.4567	0.0253	0.0075	29	0.39	135.02	0
135	Threonine (%)	0918	0.4706	0.0013	0.4567	0.0253	0.0075	29	0.55	135.00	0
135	Threonine (%)	2196	0.4710	0.0000	0.4567	0.0253	0.0075	29	0.57	135.05	0
135	Threonine (%)	0171	0.4990	0.0300	0.4567	0.0253	0.0075	29	1.67	135.00	0
135	Threonine (%)	0353	0.5000	0.0000	0.4567	0.0253	0.0075	29	1.71	135.00	0
135	Threonine (%)	2188	0.5050	0.0100	0.4567	0.0253	0.0075	29	1.91	135.05	0
135	Threonine (%)	0619	0.5150	0.0040	0.4567	0.0253	0.0075	29	2.31	135.00	0
135	Threonine (%)	0853	0.3900	0.0400	0.4567	0.0253	0.0075	29	-2.64	135.05	1
136	Tryptophan (%)	0870	0.1198	0.0066	0.1654	0.0252	0.0039	19	-1.81	136.00	0
136	Tryptophan (%)	0918	0.1400	0.0000	0.1654	0.0252	0.0039	19	-1.01	136.00	0
136	Tryptophan (%)	0504	0.1400	0.0000	0.1654	0.0252	0.0039	19	-1.01	136.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
136	Tryptophan (%)	0407	0.1405	0.0110	0.1654	0.0252	0.0039	19	-0.99	136.05	0
136	Tryptophan (%)	0910	0.1450	0.0100	0.1654	0.0252	0.0039	19	-0.81	136.00	0
136	Tryptophan (%)	0353	0.1450	0.0100	0.1654	0.0252	0.0039	19	-0.81	136.03	0
136	Tryptophan (%)	0682	0.1500	0.0000	0.1654	0.0252	0.0039	19	-0.61	136.00	0
136	Tryptophan (%)	0868	0.1570	0.0080	0.1654	0.0252	0.0039	19	-0.33	136.01	0
136	Tryptophan (%)	0098	0.1625	0.0050	0.1654	0.0252	0.0039	19	-0.11	136.02	0
136	Tryptophan (%)	2059	0.1710	0.0020	0.1654	0.0252	0.0039	19	0.22	136.03	0
136	Tryptophan (%)	0859	0.1760	0.0034	0.1654	0.0252	0.0039	19	0.42	136.03	0
136	Tryptophan (%)	0571	0.1765	0.0010	0.1654	0.0252	0.0039	19	0.44	136.03	0
136	Tryptophan (%)	0872	0.1780	0.0020	0.1654	0.0252	0.0039	19	0.50	136.03	0
136	Tryptophan (%)	0878	0.1785	0.0010	0.1654	0.0252	0.0039	19	0.52	136.01	0
136	Tryptophan (%)	0610	0.1800	0.0000	0.1654	0.0252	0.0039	19	0.58	136.99	0
136	Tryptophan (%)	0644	0.1820	0.0020	0.1654	0.0252	0.0039	19	0.66	136.01	0
136	Tryptophan (%)	0619	0.1865	0.0030	0.1654	0.0252	0.0039	19	0.84	136.00	0
136	Tryptophan (%)	0227	0.2150	0.0100	0.1654	0.0252	0.0039	19	1.97	136.00	0
136	Tryptophan (%)	2195	0.3500	0.0000	0.1654	0.0252	0.0039	19	7.34	136.05	0
137	Tyrosine (%)	0675	0.1100	0.0200	0.3677	0.0570	0.0080	23	-4.52	137.00	0
137	Tyrosine (%)	0910	0.2050	0.0100	0.3677	0.0570	0.0080	23	-2.85	137.00	0
137	Tyrosine (%)	0682	0.2670	0.0000	0.3677	0.0570	0.0080	23	-1.77	137.00	0
137	Tyrosine (%)	0626	0.3055	0.0070	0.3677	0.0570	0.0080	23	-1.09	137.05	0
137	Tyrosine (%)	0968	0.3175	0.0010	0.3677	0.0570	0.0080	23	-0.88	137.00	0
137	Tyrosine (%)	0354	0.3260	0.0040	0.3677	0.0570	0.0080	23	-0.73	137.00	0
137	Tyrosine (%)	0227	0.3400	0.0000	0.3677	0.0570	0.0080	23	-0.49	137.00	0
137	Tyrosine (%)	0407	0.3495	0.0170	0.3677	0.0570	0.0080	23	-0.32	137.05	0
137	Tyrosine (%)	0918	0.3555	0.0033	0.3677	0.0570	0.0080	23	-0.21	137.00	0
137	Tyrosine (%)	2196	0.3690	0.0000	0.3677	0.0570	0.0080	23	0.02	137.05	0
137	Tyrosine (%)	0171	0.3700	0.0060	0.3677	0.0570	0.0080	23	0.04	137.00	0
137	Tyrosine (%)	0853	0.3800	0.0200	0.3677	0.0570	0.0080	23	0.22	137.05	0
137	Tyrosine (%)	0644	0.3825	0.0050	0.3677	0.0570	0.0080	23	0.26	137.00	0
137	Tyrosine (%)	2059	0.3840	0.0040	0.3677	0.0570	0.0080	23	0.29	137.00	0
137	Tyrosine (%)	0872	0.3850	0.0100	0.3677	0.0570	0.0080	23	0.30	137.00	0
137	Tyrosine (%)	0504	0.3900	0.0000	0.3677	0.0570	0.0080	23	0.39	137.00	0
137	Tyrosine (%)	0098	0.4010	0.0020	0.3677	0.0570	0.0080	23	0.58	137.02	0
137	Tyrosine (%)	0610	0.4050	0.0100	0.3677	0.0570	0.0080	23	0.65	137.05	0
137	Tyrosine (%)	0148	0.4150	0.0100	0.3677	0.0570	0.0080	23	0.83	137.05	0
137	Tyrosine (%)	0226	0.4169	0.0241	0.3677	0.0570	0.0080	23	0.86	137.00	0
137	Tyrosine (%)	2188	0.4300	0.0200	0.3677	0.0570	0.0080	23	1.09	137.05	0
137	Tyrosine (%)	0676	0.4350	0.0100	0.3677	0.0570	0.0080	23	1.18	137.05	0
137	Tyrosine (%)	0353	0.4600	0.0000	0.3677	0.0570	0.0080	23	1.62	137.00	0
138	Valine (%)	0870	0.3787	0.0131	0.5933	0.0421	0.0116	29	-5.10	138.00	0
138	Valine (%)	0226	0.4476	0.0511	0.5933	0.0421	0.0116	29	-3.46	138.00	0
138	Valine (%)	0354	0.5020	0.0080	0.5933	0.0421	0.0116	29	-2.17	138.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
138	Valine (%)	0353	0.5200	0.0000	0.5933	0.0421	0.0116	29	-1.74	138.00	0
138	Valine (%)	0859	0.5428	0.0100	0.5933	0.0421	0.0116	29	-1.20	138.00	0
138	Valine (%)	0407	0.5430	0.0120	0.5933	0.0421	0.0116	29	-1.19	138.05	0
138	Valine (%)	0171	0.5530	0.0160	0.5933	0.0421	0.0116	29	-0.96	138.00	0
138	Valine (%)	2196	0.5690	0.0000	0.5933	0.0421	0.0116	29	-0.58	138.05	0
138	Valine (%)	0910	0.5800	0.0200	0.5933	0.0421	0.0116	29	-0.32	138.00	0
138	Valine (%)	0626	0.5855	0.0010	0.5933	0.0421	0.0116	29	-0.18	138.05	0
138	Valine (%)	0675	0.5900	0.0200	0.5933	0.0421	0.0116	29	-0.08	138.00	0
138	Valine (%)	2195	0.5900	0.0000	0.5933	0.0421	0.0116	29	-0.08	138.05	0
138	Valine (%)	0868	0.5915	0.0270	0.5933	0.0421	0.0116	29	-0.04	138.00	0
138	Valine (%)	0504	0.6000	0.0000	0.5933	0.0421	0.0116	29	0.16	138.00	0
138	Valine (%)	0619	0.6035	0.0030	0.5933	0.0421	0.0116	29	0.24	138.00	0
138	Valine (%)	0968	0.6130	0.0080	0.5933	0.0421	0.0116	29	0.47	138.00	0
138	Valine (%)	0878	0.6140	0.0100	0.5933	0.0421	0.0116	29	0.49	138.00	0
138	Valine (%)	0872	0.6150	0.0340	0.5933	0.0421	0.0116	29	0.52	138.00	0
138	Valine (%)	0676	0.6150	0.0100	0.5933	0.0421	0.0116	29	0.52	138.05	0
138	Valine (%)	2188	0.6150	0.0100	0.5933	0.0421	0.0116	29	0.52	138.05	0
138	Valine (%)	2059	0.6155	0.0010	0.5933	0.0421	0.0116	29	0.53	138.00	0
138	Valine (%)	0571	0.6185	0.0150	0.5933	0.0421	0.0116	29	0.60	138.00	0
138	Valine (%)	0644	0.6185	0.0150	0.5933	0.0421	0.0116	29	0.60	138.00	0
138	Valine (%)	0682	0.6220	0.0000	0.5933	0.0421	0.0116	29	0.68	138.00	0
138	Valine (%)	0148	0.6300	0.0000	0.5933	0.0421	0.0116	29	0.87	138.05	0
138	Valine (%)	0918	0.6345	0.0082	0.5933	0.0421	0.0116	29	0.98	138.00	0
138	Valine (%)	0610	0.6400	0.0000	0.5933	0.0421	0.0116	29	1.11	138.05	0
138	Valine (%)	0098	0.6405	0.0130	0.5933	0.0421	0.0116	29	1.12	138.02	0
138	Valine (%)	0227	0.6450	0.0300	0.5933	0.0421	0.0116	29	1.23	138.00	0
138	Valine (%)	0853	0.5200	0.0600	0.5933	0.0421	0.0116	29	-1.74	138.05	1
139	Taurine (%)	0407	0.0035	0.0010				3		139.05	0
139	Taurine (%)	0682	0.0500	0.0000				3		139.00	0
139	Taurine (%)	0504	0.1850	0.0100				3		139.00	0
139	Taurine (%)	0098	< 0.01					3		139.02	5
160	Fructose (%)	0619	0.2055	0.0030	0.4201	0.2478	0.0483	4		160.99	0
160	Fructose (%)	0227	0.2200	0.0000	0.4201	0.2478	0.0483	4		160.99	0
160	Fructose (%)	0861	0.5500	0.1000	0.4201	0.2478	0.0483	4		160.99	0
160	Fructose (%)	0148	0.7050	0.0900	0.4201	0.2478	0.0483	4		160.99	0
162	Glucose (%)	0619	0.1775	0.0030				3		162.99	0
162	Glucose (%)	0227	0.1900	0.0000				3		162.99	0
162	Glucose (%)	0148	0.3300	0.1000				3		162.99	0
162	Glucose (%)	0861	< 0.1					3		162.99	5
163	Lactose (%)	0148	0.0350	0.0100				1		163.99	0
163	Lactose (%)	0619	0.0000	0.0000				1		163.99	4
163	Lactose (%)	0861	< 0.1					1		163.99	5

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
163	Lactose (%)	0227	< 0.15					1		163.99	5
164	Maltose (%)	0148	0.1600	0.0000				2		164.99	0
164	Maltose (%)	0861	0.3500	0.1000				2		164.99	0
164	Maltose (%)	0227	< 0.15					2		164.99	5
165	Sucrose (%)	0227	1.640	0.1000	1.853	0.2638	0.1000	4		165.99	0
165	Sucrose (%)	0619	1.695	0.0300	1.853	0.2638	0.1000	4		165.99	0
165	Sucrose (%)	0861	1.850	0.1000	1.853	0.2638	0.1000	4		165.99	0
165	Sucrose (%)	0148	2.225	0.1700	1.853	0.2638	0.1000	4		165.99	0
166	Raffinose (%)	0227	0.4400	0.0000				2		166.99	0
166	Raffinose (%)	0619	0.4885	0.0210				2		166.99	0
167	Stachyose (%)	0227	0.1350	0.0100				2		167.99	0
167	Stachyose (%)	0619	0.2745	0.0130				2		167.99	0
400	Water Activity (Units)	0407	0.4555	0.0024	0.4844	0.0302	0.0039	9	-0.96	400.01	0
400	Water Activity (Units)	0843	0.4600	0.0020	0.4844	0.0302	0.0039	9	-0.81	400.99	0
400	Water Activity (Units)	0942	0.4650	0.0100	0.4844	0.0302	0.0039	9	-0.64	400.01	0
400	Water Activity (Units)	0589	0.4660	0.0080	0.4844	0.0302	0.0039	9	-0.61	400.01	0
400	Water Activity (Units)	0843	0.4815	0.0043	0.4844	0.0302	0.0039	9	-0.10	400.01	0
400	Water Activity (Units)	2009	0.4845	0.0010	0.4844	0.0302	0.0039	9	0.00	400.01	0
400	Water Activity (Units)	2073	0.4975	0.0030	0.4844	0.0302	0.0039	9	0.43	400.99	0
400	Water Activity (Units)	2181	0.5203	0.0019	0.4844	0.0302	0.0039	9	1.19	400.01	0
400	Water Activity (Units)	2109	0.5312	0.0029	0.4844	0.0302	0.0039	9	1.55	400.01	0
412	Starch, Dietary (%)	0353	11.59	0.0100				1		412.01	0
516	Arsenic, Total (mg / kg (ppm))	0045	0.9110	0.0080	1.208	0.0958	0.0621	8	-3.10	516.42	0
516	Arsenic, Total (mg / kg (ppm))	0227	1.160	0.0600	1.208	0.0958	0.0621	8	-0.50	516.53	0
516	Arsenic, Total (mg / kg (ppm))	0171	1.162	0.0030	1.208	0.0958	0.0621	8	-0.49	516.00	0
516	Arsenic, Total (mg / kg (ppm))	0186	1.195	0.0100	1.208	0.0958	0.0621	8	-0.14	516.52	0
516	Arsenic, Total (mg / kg (ppm))	0918	1.223	0.0229	1.208	0.0958	0.0621	8	0.16	516.53	0
516	Arsenic, Total (mg / kg (ppm))	0553	1.240	0.1000	1.208	0.0958	0.0621	8	0.33	516.53	0
516	Arsenic, Total (mg / kg (ppm))	0910	1.270	0.0200	1.208	0.0958	0.0621	8	0.64	516.52	0
516	Arsenic, Total (mg / kg (ppm))	0407	1.443	0.2726	1.208	0.0958	0.0621	8	2.45	516.43	0
516	Arsenic, Total (mg / kg (ppm))	0098	1.470	0.5510	1.208	0.0958	0.0621	8	2.73	516.53	1
518	Cadmium (mg / kg (ppm))	0171	0.4455	0.2090	0.7792	0.1073	0.0552	9	-3.11	518.41	0
518	Cadmium (mg / kg (ppm))	2055	0.6715	0.0330	0.7792	0.1073	0.0552	9	-1.00	518.53	0
518	Cadmium (mg / kg (ppm))	0227	0.7625	0.0070	0.7792	0.1073	0.0552	9	-0.16	518.53	0
518	Cadmium (mg / kg (ppm))	0918	0.7692	0.0325	0.7792	0.1073	0.0552	9	-0.09	518.53	0
518	Cadmium (mg / kg (ppm))	0098	0.7805	0.0430	0.7792	0.1073	0.0552	9	0.01	518.53	0
518	Cadmium (mg / kg (ppm))	0553	0.7830	0.0140	0.7792	0.1073	0.0552	9	0.04	518.53	0
518	Cadmium (mg / kg (ppm))	0186	0.8375	0.0070	0.7792	0.1073	0.0552	9	0.54	518.52	0
518	Cadmium (mg / kg (ppm))	0910	0.8500	0.1400	0.7792	0.1073	0.0552	9	0.66	518.52	0
518	Cadmium (mg / kg (ppm))	0968	1.326	0.0110	0.7792	0.1073	0.0552	9	5.09	518.43	0
520	Chromium (mg / kg (ppm))	0186	22.90	1.000	29.12	6.034	0.9824	10	-1.03	520.52	0
520	Chromium (mg / kg (ppm))	0918	23.88	0.9503	29.12	6.034	0.9824	10	-0.87	520.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
520	Chromium (mg / kg (ppm))	0968	24.08	0.4250	29.12	6.034	0.9824	10	-0.84	520.43	0
520	Chromium (mg / kg (ppm))	2055	25.07	1.805	29.12	6.034	0.9824	10	-0.67	520.53	0
520	Chromium (mg / kg (ppm))	0693	28.62	1.314	29.12	6.034	0.9824	10	-0.08	520.42	0
520	Chromium (mg / kg (ppm))	0045	29.05	0.9000	29.12	6.034	0.9824	10	-0.01	520.42	0
520	Chromium (mg / kg (ppm))	0164	31.05	1.500	29.12	6.034	0.9824	10	0.32	520.53	0
520	Chromium (mg / kg (ppm))	0553	31.90	1.200	29.12	6.034	0.9824	10	0.46	520.53	0
520	Chromium (mg / kg (ppm))	0510	36.74	0.3300	29.12	6.034	0.9824	10	1.26	520.43	0
520	Chromium (mg / kg (ppm))	0032	37.90	0.4000	29.12	6.034	0.9824	10	1.46	520.53	0
526	Lead (mg / kg (ppm))	0227	2.025	0.0900	2.201	0.1282	0.1173	7	-1.38	526.53	0
526	Lead (mg / kg (ppm))	0098	2.135	0.0700	2.201	0.1282	0.1173	7	-0.52	526.53	0
526	Lead (mg / kg (ppm))	0186	2.180	0.0600	2.201	0.1282	0.1173	7	-0.17	526.52	0
526	Lead (mg / kg (ppm))	0553	2.190	0.0200	2.201	0.1282	0.1173	7	-0.09	526.53	0
526	Lead (mg / kg (ppm))	0918	2.236	0.0513	2.201	0.1282	0.1173	7	0.27	526.53	0
526	Lead (mg / kg (ppm))	0910	2.250	0.2600	2.201	0.1282	0.1173	7	0.38	526.52	0
526	Lead (mg / kg (ppm))	0032	2.465	0.2700	2.201	0.1282	0.1173	7	2.06	526.53	0
526	Lead (mg / kg (ppm))	0171	1.133	2.114	2.201	0.1282	0.1173	7	-8.33	526.41	1
529	Mercury (µg / kg (ppb))	0186	19.00	6.000				1		529.99	0
539	Nickel (mg / kg (ppm))	0186	6.950	0.4800	12.23	3.455	0.9231	5		539.42	0
539	Nickel (mg / kg (ppm))	0186	11.25	0.7000	12.23	3.455	0.9231	5		539.52	0
539	Nickel (mg / kg (ppm))	0171	13.28	1.460	12.23	3.455	0.9231	5		539.41	0
539	Nickel (mg / kg (ppm))	0918	13.35	1.776	12.23	3.455	0.9231	5		539.53	0
539	Nickel (mg / kg (ppm))	0553	16.30	0.2000	12.23	3.455	0.9231	5		539.53	0
702	Butyric Acid (4:0) (%)	0619	0.0055	0.0010				1		702.00	0
704	Caproic Acid (6:0) (%)	0619	0.0440	0.0160				1		704.00	0
706	Caprylic acid (8:0) (%) (w/w)	0619	0.0050	0.0020				1		706.01	0
708	Capric acid (10:0) (%) (w/w)	0619	0.0005	0.0010				0		708.01	4
710	Lauric Acid (12:0) (%) (w/w)	0226	0.0020	0.0003				3		710.99	0
710	Lauric Acid (12:0) (%) (w/w)	0676	0.0075	0.0010				3		710.99	0
710	Lauric Acid (12:0) (%) (w/w)	0619	0.0320	0.0280				3		710.01	0
710	Lauric Acid (12:0) (%) (w/w)	0164	< 0.02					3		710.99	5
714	Myristic Acid (14:0) (%) (w/w)	0226	0.0060	0.0003				2		714.99	0
714	Myristic Acid (14:0) (%) (w/w)	0619	0.0365	0.0310				2		714.01	0
714	Myristic Acid (14:0) (%) (w/w)	0164	< 0.02					2		714.99	5
716	Palmitic Acid (16:0) (%) (w/w)	0226	0.4317	0.0025				3		716.99	0
716	Palmitic Acid (16:0) (%) (w/w)	0619	0.5425	0.0590				3		716.01	0
716	Palmitic Acid (16:0) (%) (w/w)	0164	0.5950	0.1300				3		716.99	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0676	0.0070	0.0000				2		718.99	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0226	0.0120	0.0048				2		718.99	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0619	0.0000	0.0000				2		718.01	4
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0164	< 0.02					2		718.99	5
722	Stearic Acid (18:0) (%) (w/w)	0226	0.0494	0.0005				2		722.99	0
722	Stearic Acid (18:0) (%) (w/w)	0619	0.0580	0.0260				2		722.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
724	Oleic Acid (9c-18:1) (% (w/w))	0226	0.3215	0.0043				3	724.99	0	
724	Oleic Acid (9c-18:1) (% (w/w))	0619	0.3560	0.0140				3	724.01	0	
724	Oleic Acid (9c-18:1) (% (w/w))	0164	0.5200	0.0800				3	724.99	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0148	0.4200	0.0000	0.4824	0.0515	0.0331	5	726.02	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0226	0.4437	0.0183	0.4824	0.0515	0.0331	5	726.99	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0676	0.4870	0.0060	0.4824	0.0515	0.0331	5	726.99	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0164	0.5150	0.0700	0.4824	0.0515	0.0331	5	726.99	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0619	0.5465	0.0710	0.4824	0.0515	0.0331	5	726.01	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v	0226	0.0497	0.0021				3	728.99	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v	0164	0.0550	0.0100				3	728.99	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v	0676	0.0625	0.0010				3	728.99	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v	0619	0.1025	0.0430				3	728.01	1	
730	Arachidic Acid (20:0) (% (w/w))	0226	0.0108	0.0002				2	730.99	0	
730	Arachidic Acid (20:0) (% (w/w))	0619	0.0425	0.0210				2	730.01	0	
732	Gondoic Acid (11c-20:1) (% (w/w))	0226	0.0109	0.0004				2	732.99	0	
732	Gondoic Acid (11c-20:1) (% (w/w))	0619	0.0375	0.0130				2	732.01	0	
736	Arachidonic Acid (5c,8c,11c,14c-20:4) (% (w/v	0676	0.0120	0.0000				1	736.99	0	
736	Arachidonic Acid (5c,8c,11c,14c-20:4) (% (w/v	0619	0.0000	0.0000				1	736.01	4	
738	Mead Acid (11c,14c,17c-20:3) (% (w/w))	0619	0.0000	0.0000				0	738.01	4	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c	0619	0.0000	0.0000				0	740.01	4	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c	0226	< 0.005					0	740.99	5	
742	Behenic Acid (22:0) (% (w/w))	0226	0.0123	0.0018				1	742.99	0	
742	Behenic Acid (22:0) (% (w/w))	0164	< 0.02					1	742.99	5	
744	Erucic Acid (13c-22:1) (% (w/w))	0619	0.0000	0.0000				0	744.01	4	
744	Erucic Acid (13c-22:1) (% (w/w))	0226	< 0.005					0	744.99	5	
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,1	0619	0.0000	0.0000				0	746.01	4	
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,1	0226	< 0.005					0	746.99	5	
748	Lignoceric Acid (24:0) (% (w/w))	0226	0.0113	0.0007				1	748.99	0	
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16	0619	0.0000	0.0000				0	750.01	4	
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16	0226	< 0.005					0	750.99	5	
752	Nervonic Acid (24:1) isomers (% (w/w))	0619	0.0000	0.0000				0	752.01	4	
752	Nervonic Acid (24:1) isomers (% (w/w))	0226	< 0.005					0	752.99	5	
754	Total n-3 Polyunsaturated (Omega-3) Fatty Aci	0148	0.0600	0.0000				3	754.02	0	
754	Total n-3 Polyunsaturated (Omega-3) Fatty Aci	0676	0.0600	0.0000				3	754.99	0	
754	Total n-3 Polyunsaturated (Omega-3) Fatty Aci	0164	0.0650	0.0100				3	754.99	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Aci	0148	0.4200	0.0000				3	756.01	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Aci	0164	0.4950	0.0700				3	756.99	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Aci	0676	0.5000	0.0000				3	756.99	0	
758	Total Saturated Fatty Acids (% (w/w))	0164	0.7800	0.1600				1	758.99	0	
762	Total Monounsaturated Fatty Acids (% (w/w))	0164	0.5100	0.1000				1	762.99	0	
766	Total Polyunsaturated Fatty Acids (% (w/w))	0164	0.5750	0.0900				1	766.99	0	
770	Total Fat (equivalent to NLEA) (% (w/w))	0164	2.025	0.3500				1	770.99	0	

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
772	Total Fatty Acids (% (w/w))	0226	1.361	0.0144				2	772.99	0	
772	Total Fatty Acids (% (w/w))	0164	1.935	0.3300				2	772.99	0	

Notes: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2 and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = Rejected for duplicates too far apart, 2 = Rejected as outlier, 5 = A reporting limit and 4 = zeros submitted as values. Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). Flag 3 indicates not used in statistics.