

Animal Feed Scheme
Dry Dog Food

Labs Reporting: 217
Analytes Reported 103
Issue Date : 04/30/2019

Test Material Code # 201923

Analyte Proficiency Testing Report

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
000	Urea (%)	2246	0.0200	0.0200				2		000.02	0
000	Urea (%)	0278	0.1000	0.0000				2		000.02	0
001	Loss on Drying (%)	0505	6.385	0.4100	7.319	0.2325	0.1003	88	-4.02	001.07	0
001	Loss on Drying (%)	2155	6.430	0.2200	7.319	0.2325	0.1003	88	-3.83	001.99	0
001	Loss on Drying (%)	2081	6.595	0.1700	7.319	0.2325	0.1003	88	-3.12	001.99	0
001	Loss on Drying (%)	0845	6.615	0.4700	7.319	0.2325	0.1003	88	-3.03	001.07	0
001	Loss on Drying (%)	0889	6.730	0.0200	7.319	0.2325	0.1003	88	-2.54	001.99	0
001	Loss on Drying (%)	0345	6.810	0.0200	7.319	0.2325	0.1003	88	-2.19	001.07	0
001	Loss on Drying (%)	0074	6.985	0.0300	7.319	0.2325	0.1003	88	-1.44	001.07	0
001	Loss on Drying (%)	0019	7.000	0.2000	7.319	0.2325	0.1003	88	-1.37	001.07	0
001	Loss on Drying (%)	0609	7.000	0.0000	7.319	0.2325	0.1003	88	-1.37	001.07	0
001	Loss on Drying (%)	2190	7.010	0.0800	7.319	0.2325	0.1003	88	-1.33	001.07	0
001	Loss on Drying (%)	0034	7.015	0.0500	7.319	0.2325	0.1003	88	-1.31	001.07	0
001	Loss on Drying (%)	0876	7.016	0.1474	7.319	0.2325	0.1003	88	-1.31	001.07	0
001	Loss on Drying (%)	0038	7.035	0.1900	7.319	0.2325	0.1003	88	-1.22	001.07	0
001	Loss on Drying (%)	0676	7.055	0.0100	7.319	0.2325	0.1003	88	-1.14	001.99	0
001	Loss on Drying (%)	0187	7.060	0.0200	7.319	0.2325	0.1003	88	-1.12	001.07	0
001	Loss on Drying (%)	0683	7.070	0.0600	7.319	0.2325	0.1003	88	-1.07	001.07	0
001	Loss on Drying (%)	0504	7.090	0.0600	7.319	0.2325	0.1003	88	-0.99	001.00	0
001	Loss on Drying (%)	2264	7.109	0.3903	7.319	0.2325	0.1003	88	-0.90	001.07	0
001	Loss on Drying (%)	0590	7.110	0.1000	7.319	0.2325	0.1003	88	-0.90	001.07	0
001	Loss on Drying (%)	2156	7.120	0.0800	7.319	0.2325	0.1003	88	-0.86	001.07	0
001	Loss on Drying (%)	0610	7.125	0.0100	7.319	0.2325	0.1003	88	-0.84	001.05	0
001	Loss on Drying (%)	2146	7.150	0.0600	7.319	0.2325	0.1003	88	-0.73	001.07	0
001	Loss on Drying (%)	0186	7.160	0.0400	7.319	0.2325	0.1003	88	-0.69	001.99	0
001	Loss on Drying (%)	0139	7.180	0.1600	7.319	0.2325	0.1003	88	-0.60	001.07	0
001	Loss on Drying (%)	0948	7.185	0.0500	7.319	0.2325	0.1003	88	-0.58	001.99	0
001	Loss on Drying (%)	2009	7.194	0.0064	7.319	0.2325	0.1003	88	-0.54	001.07	0
001	Loss on Drying (%)	2222	7.209	0.0570	7.319	0.2325	0.1003	88	-0.48	001.07	0
001	Loss on Drying (%)	2115	7.210	0.0200	7.319	0.2325	0.1003	88	-0.47	001.99	0
001	Loss on Drying (%)	2259	7.225	0.1700	7.319	0.2325	0.1003	88	-0.41	001.07	0
001	Loss on Drying (%)	2126	7.240	0.0000	7.319	0.2325	0.1003	88	-0.34	001.99	0
001	Loss on Drying (%)	0689	7.250	0.1000	7.319	0.2325	0.1003	88	-0.30	001.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
001	Loss on Drying (%)	0169	7.265	0.0100	7.319	0.2325	0.1003	88	-0.23	001.00	0
001	Loss on Drying (%)	0309	7.277	0.0550	7.319	0.2325	0.1003	88	-0.18	001.00	0
001	Loss on Drying (%)	2219	7.280	0.1200	7.319	0.2325	0.1003	88	-0.17	001.07	0
001	Loss on Drying (%)	0656	7.280	0.0600	7.319	0.2325	0.1003	88	-0.17	001.99	0
001	Loss on Drying (%)	0083	7.310	0.0800	7.319	0.2325	0.1003	88	-0.04	001.07	0
001	Loss on Drying (%)	2124	7.310	0.1800	7.319	0.2325	0.1003	88	-0.04	001.03	0
001	Loss on Drying (%)	0644	7.320	0.0000	7.319	0.2325	0.1003	88	0.00	001.05	0
001	Loss on Drying (%)	0027	7.323	0.0110	7.319	0.2325	0.1003	88	0.01	001.00	0
001	Loss on Drying (%)	2059	7.325	0.1300	7.319	0.2325	0.1003	88	0.02	001.07	0
001	Loss on Drying (%)	2192	7.330	0.1200	7.319	0.2325	0.1003	88	0.05	001.07	0
001	Loss on Drying (%)	2018	7.330	0.0400	7.319	0.2325	0.1003	88	0.05	001.99	0
001	Loss on Drying (%)	0003	7.350	0.0000	7.319	0.2325	0.1003	88	0.13	001.07	0
001	Loss on Drying (%)	0045	7.350	0.0800	7.319	0.2325	0.1003	88	0.13	001.07	0
001	Loss on Drying (%)	0035	7.355	0.0300	7.319	0.2325	0.1003	88	0.15	001.07	0
001	Loss on Drying (%)	0226	7.355	0.0700	7.319	0.2325	0.1003	88	0.15	001.07	0
001	Loss on Drying (%)	0675	7.355	0.0500	7.319	0.2325	0.1003	88	0.15	001.07	0
001	Loss on Drying (%)	0918	7.355	0.1100	7.319	0.2325	0.1003	88	0.15	001.99	0
001	Loss on Drying (%)	2246	7.355	0.0100	7.319	0.2325	0.1003	88	0.15	001.99	0
001	Loss on Drying (%)	0581	7.375	0.0900	7.319	0.2325	0.1003	88	0.24	001.07	0
001	Loss on Drying (%)	0142	7.380	0.0400	7.319	0.2325	0.1003	88	0.26	001.07	0
001	Loss on Drying (%)	0278	7.385	0.0500	7.319	0.2325	0.1003	88	0.28	001.07	0
001	Loss on Drying (%)	0065	7.390	0.0600	7.319	0.2325	0.1003	88	0.30	001.07	0
001	Loss on Drying (%)	0619	7.400	0.0600	7.319	0.2325	0.1003	88	0.35	001.03	0
001	Loss on Drying (%)	0357	7.404	0.0078	7.319	0.2325	0.1003	88	0.36	001.99	0
001	Loss on Drying (%)	0171	7.405	0.0700	7.319	0.2325	0.1003	88	0.37	001.07	0
001	Loss on Drying (%)	2012	7.410	0.1200	7.319	0.2325	0.1003	88	0.39	001.07	0
001	Loss on Drying (%)	0571	7.415	0.0300	7.319	0.2325	0.1003	88	0.41	001.07	0
001	Loss on Drying (%)	0868	7.420	0.0000	7.319	0.2325	0.1003	88	0.43	001.03	0
001	Loss on Drying (%)	0089	7.430	0.0000	7.319	0.2325	0.1003	88	0.48	001.07	0
001	Loss on Drying (%)	0066	7.435	0.1300	7.319	0.2325	0.1003	88	0.50	001.07	0
001	Loss on Drying (%)	0366	7.450	0.3000	7.319	0.2325	0.1003	88	0.56	001.07	0
001	Loss on Drying (%)	2187	7.450	0.0400	7.319	0.2325	0.1003	88	0.56	001.07	0
001	Loss on Drying (%)	0849	7.455	0.0100	7.319	0.2325	0.1003	88	0.58	001.07	0
001	Loss on Drying (%)	2053	7.455	0.0300	7.319	0.2325	0.1003	88	0.58	001.07	0
001	Loss on Drying (%)	0049	7.460	0.0800	7.319	0.2325	0.1003	88	0.60	001.07	0
001	Loss on Drying (%)	0682	7.460	0.0000	7.319	0.2325	0.1003	88	0.60	001.07	0
001	Loss on Drying (%)	0695	7.482	0.0130	7.319	0.2325	0.1003	88	0.70	001.07	0
001	Loss on Drying (%)	0853	7.500	0.0000	7.319	0.2325	0.1003	88	0.78	001.99	0
001	Loss on Drying (%)	2252	7.505	0.3700	7.319	0.2325	0.1003	88	0.80	001.07	0
001	Loss on Drying (%)	0353	7.510	0.0600	7.319	0.2325	0.1003	88	0.82	001.07	0
001	Loss on Drying (%)	2220	7.510	0.0800	7.319	0.2325	0.1003	88	0.82	001.07	0
001	Loss on Drying (%)	0662	7.516	0.3906	7.319	0.2325	0.1003	88	0.85	001.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
001	Loss on Drying (%)	0560	7.535	0.4900	7.319	0.2325	0.1003	88	0.93	001.99	0
001	Loss on Drying (%)	0878	7.545	0.0700	7.319	0.2325	0.1003	88	0.97	001.03	0
001	Loss on Drying (%)	0910	7.550	0.1000	7.319	0.2325	0.1003	88	0.99	001.99	0
001	Loss on Drying (%)	0852	7.550	0.0400	7.319	0.2325	0.1003	88	0.99	001.99	0
001	Loss on Drying (%)	0037	7.560	0.0800	7.319	0.2325	0.1003	88	1.03	001.99	0
001	Loss on Drying (%)	0893	7.560	0.0200	7.319	0.2325	0.1003	88	1.03	001.03	0
001	Loss on Drying (%)	0638	7.570	0.0200	7.319	0.2325	0.1003	88	1.08	001.07	0
001	Loss on Drying (%)	0872	7.585	0.0700	7.319	0.2325	0.1003	88	1.14	001.07	0
001	Loss on Drying (%)	0098	7.595	0.0700	7.319	0.2325	0.1003	88	1.19	001.07	0
001	Loss on Drying (%)	0843	7.675	0.5500	7.319	0.2325	0.1003	88	1.53	001.07	0
001	Loss on Drying (%)	0407	7.737	0.1209	7.319	0.2325	0.1003	88	1.80	001.07	0
001	Loss on Drying (%)	0693	7.771	0.4590	7.319	0.2325	0.1003	88	1.94	001.07	0
001	Loss on Drying (%)	2109	7.860	0.0400	7.319	0.2325	0.1003	88	2.33	001.07	0
001	Loss on Drying (%)	0529	7.925	0.0300	7.319	0.2325	0.1003	88	2.60	001.07	0
001	Loss on Drying (%)	0873	9.300	0.0800	7.319	0.2325	0.1003	88	8.52	001.07	0
001	Loss on Drying (%)	0613	6.880	1.140	7.319	0.2325	0.1003	88	-1.89	001.07	1
001	Loss on Drying (%)	0510	92.55	0.1000	7.319	0.2325	0.1003	88	366.64	001.99	2
002	Protein, Crude (%)	0011	21.30	0.0000	24.39	0.3022	0.1467	217	-10.22	002.11	0
002	Protein, Crude (%)	0852	22.65	0.1000	24.39	0.3022	0.1467	217	-5.75	002.05	0
002	Protein, Crude (%)	0889	22.82	0.0100	24.39	0.3022	0.1467	217	-5.21	002.11	0
002	Protein, Crude (%)	2259	23.39	0.2400	24.39	0.3022	0.1467	217	-3.30	002.01	0
002	Protein, Crude (%)	0910	23.40	0.0000	24.39	0.3022	0.1467	217	-3.27	002.01	0
002	Protein, Crude (%)	0643	23.47	0.0900	24.39	0.3022	0.1467	217	-3.06	002.99	0
002	Protein, Crude (%)	0874	23.51	0.0200	24.39	0.3022	0.1467	217	-2.91	002.04	0
002	Protein, Crude (%)	0646	23.54	0.2400	24.39	0.3022	0.1467	217	-2.81	002.06	0
002	Protein, Crude (%)	0651	23.55	0.0282	24.39	0.3022	0.1467	217	-2.79	002.05	0
002	Protein, Crude (%)	0921	23.59	0.0000	24.39	0.3022	0.1467	217	-2.64	002.06	0
002	Protein, Crude (%)	0613	23.63	0.0600	24.39	0.3022	0.1467	217	-2.51	002.11	0
002	Protein, Crude (%)	2192	23.69	0.2600	24.39	0.3022	0.1467	217	-2.31	002.05	0
002	Protein, Crude (%)	0860	23.82	0.0000	24.39	0.3022	0.1467	217	-1.89	002.01	0
002	Protein, Crude (%)	2190	23.87	0.2500	24.39	0.3022	0.1467	217	-1.73	002.05	0
002	Protein, Crude (%)	0650	23.87	0.0800	24.39	0.3022	0.1467	217	-1.71	002.06	0
002	Protein, Crude (%)	0921	23.87	0.0000	24.39	0.3022	0.1467	217	-1.71	002.05	0
002	Protein, Crude (%)	0294	23.90	0.0000	24.39	0.3022	0.1467	217	-1.62	002.06	0
002	Protein, Crude (%)	0683	23.93	0.1700	24.39	0.3022	0.1467	217	-1.53	002.05	0
002	Protein, Crude (%)	0510	23.95	0.1000	24.39	0.3022	0.1467	217	-1.45	002.06	0
002	Protein, Crude (%)	0941	23.96	0.3600	24.39	0.3022	0.1467	217	-1.42	002.05	0
002	Protein, Crude (%)	0662	23.97	0.0625	24.39	0.3022	0.1467	217	-1.39	002.01	0
002	Protein, Crude (%)	0644	23.98	0.2400	24.39	0.3022	0.1467	217	-1.36	002.06	0
002	Protein, Crude (%)	0695	23.99	0.2880	24.39	0.3022	0.1467	217	-1.33	002.06	0
002	Protein, Crude (%)	0653	24.00	0.1200	24.39	0.3022	0.1467	217	-1.28	002.01	0
002	Protein, Crude (%)	0955	24.00	0.2000	24.39	0.3022	0.1467	217	-1.28	002.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein, Crude (%)	0849	24.01	0.0720	24.39	0.3022	0.1467	217	-1.26	002.05	0
002	Protein, Crude (%)	0598	24.02	0.5300	24.39	0.3022	0.1467	217	-1.23	002.06	0
002	Protein, Crude (%)	2190	24.02	0.1100	24.39	0.3022	0.1467	217	-1.23	002.06	0
002	Protein, Crude (%)	0852	24.03	0.0200	24.39	0.3022	0.1467	217	-1.19	002.11	0
002	Protein, Crude (%)	0968	24.04	0.2300	24.39	0.3022	0.1467	217	-1.16	002.01	0
002	Protein, Crude (%)	2006	24.05	0.3800	24.39	0.3022	0.1467	217	-1.12	002.05	0
002	Protein, Crude (%)	0619	24.05	0.1000	24.39	0.3022	0.1467	217	-1.12	002.01	0
002	Protein, Crude (%)	0619	24.05	0.1000	24.39	0.3022	0.1467	217	-1.12	002.05	0
002	Protein, Crude (%)	2022	24.05	0.1400	24.39	0.3022	0.1467	217	-1.12	002.05	0
002	Protein, Crude (%)	2155	24.07	0.4600	24.39	0.3022	0.1467	217	-1.06	002.06	0
002	Protein, Crude (%)	0354	24.08	0.0100	24.39	0.3022	0.1467	217	-1.04	002.05	0
002	Protein, Crude (%)	2073	24.08	0.1000	24.39	0.3022	0.1467	217	-1.02	002.05	0
002	Protein, Crude (%)	2144	24.08	0.2000	24.39	0.3022	0.1467	217	-1.02	002.05	0
002	Protein, Crude (%)	0941	24.08	0.3800	24.39	0.3022	0.1467	217	-1.02	002.06	0
002	Protein, Crude (%)	0956	24.10	0.0000	24.39	0.3022	0.1467	217	-0.95	002.06	0
002	Protein, Crude (%)	2169	24.10	0.2000	24.39	0.3022	0.1467	217	-0.95	002.06	0
002	Protein, Crude (%)	2181	24.12	0.0700	24.39	0.3022	0.1467	217	-0.90	002.05	0
002	Protein, Crude (%)	0001	24.12	0.0500	24.39	0.3022	0.1467	217	-0.90	002.06	0
002	Protein, Crude (%)	0119	24.12	0.2130	24.39	0.3022	0.1467	217	-0.89	002.06	0
002	Protein, Crude (%)	0089	24.13	0.0000	24.39	0.3022	0.1467	217	-0.85	002.06	0
002	Protein, Crude (%)	0652	24.15	0.1000	24.39	0.3022	0.1467	217	-0.79	002.01	0
002	Protein, Crude (%)	0011	24.15	0.1000	24.39	0.3022	0.1467	217	-0.79	002.06	0
002	Protein, Crude (%)	0098	24.15	0.0600	24.39	0.3022	0.1467	217	-0.79	002.08	0
002	Protein, Crude (%)	0939	24.16	0.3900	24.39	0.3022	0.1467	217	-0.77	002.05	0
002	Protein, Crude (%)	0003	24.16	0.0900	24.39	0.3022	0.1467	217	-0.77	002.06	0
002	Protein, Crude (%)	0202	24.16	0.0370	24.39	0.3022	0.1467	217	-0.75	002.06	0
002	Protein, Crude (%)	0042	24.17	0.1400	24.39	0.3022	0.1467	217	-0.72	002.02	0
002	Protein, Crude (%)	2196	24.18	0.0000	24.39	0.3022	0.1467	217	-0.69	002.01	0
002	Protein, Crude (%)	2188	24.19	0.3100	24.39	0.3022	0.1467	217	-0.67	002.01	0
002	Protein, Crude (%)	0885	24.19	0.1900	24.39	0.3022	0.1467	217	-0.65	002.05	0
002	Protein, Crude (%)	0893	24.20	0.0500	24.39	0.3022	0.1467	217	-0.64	002.05	0
002	Protein, Crude (%)	0689	24.20	0.0000	24.39	0.3022	0.1467	217	-0.62	002.05	0
002	Protein, Crude (%)	0675	24.21	0.0800	24.39	0.3022	0.1467	217	-0.59	002.05	0
002	Protein, Crude (%)	0036	24.21	0.0793	24.39	0.3022	0.1467	217	-0.59	002.02	0
002	Protein, Crude (%)	2081	24.23	0.3300	24.39	0.3022	0.1467	217	-0.54	002.06	0
002	Protein, Crude (%)	0897	24.23	0.0000	24.39	0.3022	0.1467	217	-0.52	002.05	0
002	Protein, Crude (%)	0592	24.24	0.0700	24.39	0.3022	0.1467	217	-0.51	002.06	0
002	Protein, Crude (%)	0939	24.24	0.1400	24.39	0.3022	0.1467	217	-0.49	002.06	0
002	Protein, Crude (%)	0656	24.25	0.2800	24.39	0.3022	0.1467	217	-0.46	002.01	0
002	Protein, Crude (%)	0948	24.26	0.0600	24.39	0.3022	0.1467	217	-0.42	002.05	0
002	Protein, Crude (%)	0881	24.26	0.0550	24.39	0.3022	0.1467	217	-0.42	002.05	0
002	Protein, Crude (%)	0546	24.26	0.0240	24.39	0.3022	0.1467	217	-0.42	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein, Crude (%)	2187	24.27	0.0500	24.39	0.3022	0.1467	217	-0.41	002.06	0
002	Protein, Crude (%)	0009	24.27	0.0200	24.39	0.3022	0.1467	217	-0.39	002.06	0
002	Protein, Crude (%)	2264	24.28	0.1840	24.39	0.3022	0.1467	217	-0.36	002.05	0
002	Protein, Crude (%)	0870	24.28	0.1030	24.39	0.3022	0.1467	217	-0.35	002.01	0
002	Protein, Crude (%)	0169	24.29	0.0100	24.39	0.3022	0.1467	217	-0.34	002.02	0
002	Protein, Crude (%)	2012	24.29	0.1900	24.39	0.3022	0.1467	217	-0.34	002.05	0
002	Protein, Crude (%)	0541	24.30	0.4100	24.39	0.3022	0.1467	217	-0.31	002.06	0
002	Protein, Crude (%)	2246	24.30	0.0400	24.39	0.3022	0.1467	217	-0.29	002.05	0
002	Protein, Crude (%)	0955	24.30	0.6000	24.39	0.3022	0.1467	217	-0.29	002.06	0
002	Protein, Crude (%)	0032	24.31	0.1200	24.39	0.3022	0.1467	217	-0.27	002.06	0
002	Protein, Crude (%)	2058	24.31	0.0400	24.39	0.3022	0.1467	217	-0.26	002.04	0
002	Protein, Crude (%)	2124	24.31	0.1000	24.39	0.3022	0.1467	217	-0.26	002.06	0
002	Protein, Crude (%)	2222	24.31	0.1600	24.39	0.3022	0.1467	217	-0.26	002.06	0
002	Protein, Crude (%)	0405	24.31	0.0000	24.39	0.3022	0.1467	217	-0.25	002.04	0
002	Protein, Crude (%)	2128	24.31	0.2270	24.39	0.3022	0.1467	217	-0.25	002.03	0
002	Protein, Crude (%)	0529	24.32	0.1000	24.39	0.3022	0.1467	217	-0.23	002.06	0
002	Protein, Crude (%)	0868	24.32	0.0200	24.39	0.3022	0.1467	217	-0.23	002.06	0
002	Protein, Crude (%)	0298	24.33	0.0200	24.39	0.3022	0.1467	217	-0.19	002.06	0
002	Protein, Crude (%)	2162	24.33	0.2800	24.39	0.3022	0.1467	217	-0.19	002.06	0
002	Protein, Crude (%)	0100	24.34	0.3300	24.39	0.3022	0.1467	217	-0.18	002.06	0
002	Protein, Crude (%)	2129	24.34	0.0100	24.39	0.3022	0.1467	217	-0.18	002.06	0
002	Protein, Crude (%)	2009	24.34	0.0024	24.39	0.3022	0.1467	217	-0.15	002.05	0
002	Protein, Crude (%)	0357	24.34	0.3125	24.39	0.3022	0.1467	217	-0.15	002.06	0
002	Protein, Crude (%)	0098	24.35	0.1000	24.39	0.3022	0.1467	217	-0.13	002.06	0
002	Protein, Crude (%)	0278	24.35	0.1000	24.39	0.3022	0.1467	217	-0.13	002.06	0
002	Protein, Crude (%)	0682	24.35	0.0000	24.39	0.3022	0.1467	217	-0.13	002.06	0
002	Protein, Crude (%)	0066	24.36	0.1300	24.39	0.3022	0.1467	217	-0.11	002.02	0
002	Protein, Crude (%)	0638	24.36	0.0100	24.39	0.3022	0.1467	217	-0.11	002.04	0
002	Protein, Crude (%)	0171	24.36	0.1200	24.39	0.3022	0.1467	217	-0.09	002.06	0
002	Protein, Crude (%)	2198	24.37	0.0100	24.39	0.3022	0.1467	217	-0.08	002.01	0
002	Protein, Crude (%)	2109	24.37	0.0100	24.39	0.3022	0.1467	217	-0.08	002.05	0
002	Protein, Crude (%)	2058	24.37	0.0300	24.39	0.3022	0.1467	217	-0.08	002.06	0
002	Protein, Crude (%)	2168	24.37	0.0500	24.39	0.3022	0.1467	217	-0.08	002.06	0
002	Protein, Crude (%)	0065	24.37	0.1150	24.39	0.3022	0.1467	217	-0.06	002.06	0
002	Protein, Crude (%)	0038	24.38	0.0600	24.39	0.3022	0.1467	217	-0.02	002.06	0
002	Protein, Crude (%)	0868	24.39	0.0500	24.39	0.3022	0.1467	217	-0.01	002.05	0
002	Protein, Crude (%)	0028	24.39	0.5300	24.39	0.3022	0.1467	217	-0.01	002.06	0
002	Protein, Crude (%)	0353	24.39	0.0300	24.39	0.3022	0.1467	217	-0.01	002.06	0
002	Protein, Crude (%)	0843	24.39	0.0200	24.39	0.3022	0.1467	217	0.01	002.06	0
002	Protein, Crude (%)	0505	24.40	0.0500	24.39	0.3022	0.1467	217	0.02	002.06	0
002	Protein, Crude (%)	0948	24.40	0.0000	24.39	0.3022	0.1467	217	0.03	002.06	0
002	Protein, Crude (%)	0674	24.40	0.0000	24.39	0.3022	0.1467	217	0.04	002.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein, Crude (%)	0354	24.40	0.0200	24.39	0.3022	0.1467	217	0.04	002.06	0
002	Protein, Crude (%)	0619	24.40	0.0000	24.39	0.3022	0.1467	217	0.04	002.06	0
002	Protein, Crude (%)	0957	24.41	0.0900	24.39	0.3022	0.1467	217	0.06	002.06	0
002	Protein, Crude (%)	0164	24.41	0.0600	24.39	0.3022	0.1467	217	0.07	002.06	0
002	Protein, Crude (%)	0345	24.41	0.0200	24.39	0.3022	0.1467	217	0.07	002.06	0
002	Protein, Crude (%)	0848	24.42	0.0100	24.39	0.3022	0.1467	217	0.09	002.06	0
002	Protein, Crude (%)	0504	24.42	0.2800	24.39	0.3022	0.1467	217	0.11	002.04	0
002	Protein, Crude (%)	2220	24.42	0.1600	24.39	0.3022	0.1467	217	0.11	002.06	0
002	Protein, Crude (%)	0968	24.42	0.0040	24.39	0.3022	0.1467	217	0.12	002.06	0
002	Protein, Crude (%)	0958	24.43	0.1300	24.39	0.3022	0.1467	217	0.12	002.06	0
002	Protein, Crude (%)	2018	24.43	0.0200	24.39	0.3022	0.1467	217	0.14	002.06	0
002	Protein, Crude (%)	0226	24.44	0.0700	24.39	0.3022	0.1467	217	0.16	002.06	0
002	Protein, Crude (%)	0626	24.44	0.0160	24.39	0.3022	0.1467	217	0.16	002.06	0
002	Protein, Crude (%)	0037	24.44	0.3400	24.39	0.3022	0.1467	217	0.17	002.06	0
002	Protein, Crude (%)	0205	24.45	0.0300	24.39	0.3022	0.1467	217	0.19	002.06	0
002	Protein, Crude (%)	0083	24.45	0.2200	24.39	0.3022	0.1467	217	0.20	002.06	0
002	Protein, Crude (%)	0045	24.45	0.3000	24.39	0.3022	0.1467	217	0.20	002.06	0
002	Protein, Crude (%)	2184	24.46	0.0000	24.39	0.3022	0.1467	217	0.24	002.06	0
002	Protein, Crude (%)	2150	24.47	0.0100	24.39	0.3022	0.1467	217	0.25	002.06	0
002	Protein, Crude (%)	0878	24.48	0.0500	24.39	0.3022	0.1467	217	0.29	002.05	0
002	Protein, Crude (%)	0511	24.48	0.3100	24.39	0.3022	0.1467	217	0.29	002.06	0
002	Protein, Crude (%)	0676	24.48	0.1900	24.39	0.3022	0.1467	217	0.29	002.06	0
002	Protein, Crude (%)	0142	24.48	0.0400	24.39	0.3022	0.1467	217	0.30	002.06	0
002	Protein, Crude (%)	0964	24.48	0.0340	24.39	0.3022	0.1467	217	0.30	002.06	0
002	Protein, Crude (%)	0845	24.49	0.0700	24.39	0.3022	0.1467	217	0.32	002.00	0
002	Protein, Crude (%)	2115	24.49	0.2700	24.39	0.3022	0.1467	217	0.32	002.05	0
002	Protein, Crude (%)	0042	24.49	0.1000	24.39	0.3022	0.1467	217	0.34	002.06	0
002	Protein, Crude (%)	2193	24.50	0.2700	24.39	0.3022	0.1467	217	0.35	002.01	0
002	Protein, Crude (%)	0019	24.50	0.0100	24.39	0.3022	0.1467	217	0.35	002.06	0
002	Protein, Crude (%)	0017	24.50	0.0000	24.39	0.3022	0.1467	217	0.37	002.06	0
002	Protein, Crude (%)	0051	24.50	0.0000	24.39	0.3022	0.1467	217	0.37	002.06	0
002	Protein, Crude (%)	0366	24.50	0.4000	24.39	0.3022	0.1467	217	0.37	002.06	0
002	Protein, Crude (%)	0571	24.51	0.5700	24.39	0.3022	0.1467	217	0.39	002.06	0
002	Protein, Crude (%)	0589	24.51	0.3100	24.39	0.3022	0.1467	217	0.39	002.06	0
002	Protein, Crude (%)	0144	24.51	0.0600	24.39	0.3022	0.1467	217	0.40	002.06	0
002	Protein, Crude (%)	2053	24.51	0.3600	24.39	0.3022	0.1467	217	0.40	002.06	0
002	Protein, Crude (%)	0613	24.52	0.0500	24.39	0.3022	0.1467	217	0.42	002.06	0
002	Protein, Crude (%)	2252	24.52	0.0900	24.39	0.3022	0.1467	217	0.42	002.06	0
002	Protein, Crude (%)	0074	24.52	0.2000	24.39	0.3022	0.1467	217	0.44	002.06	0
002	Protein, Crude (%)	0263	24.53	0.0100	24.39	0.3022	0.1467	217	0.46	002.06	0
002	Protein, Crude (%)	0889	24.53	0.0800	24.39	0.3022	0.1467	217	0.47	002.01	0
002	Protein, Crude (%)	0027	24.53	0.3450	24.39	0.3022	0.1467	217	0.48	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein, Crude (%)	0693	24.53	0.2880	24.39	0.3022	0.1467	217	0.48	002.06	0
002	Protein, Crude (%)	0049	24.55	0.3500	24.39	0.3022	0.1467	217	0.52	002.06	0
002	Protein, Crude (%)	0590	24.55	0.2200	24.39	0.3022	0.1467	217	0.54	002.06	0
002	Protein, Crude (%)	0504	24.56	0.2500	24.39	0.3022	0.1467	217	0.55	002.06	0
002	Protein, Crude (%)	2126	24.57	0.1700	24.39	0.3022	0.1467	217	0.59	002.06	0
002	Protein, Crude (%)	0870	24.57	0.0019	24.39	0.3022	0.1467	217	0.60	002.06	0
002	Protein, Crude (%)	0108	24.57	0.0400	24.39	0.3022	0.1467	217	0.60	002.06	0
002	Protein, Crude (%)	0726	24.57	0.0400	24.39	0.3022	0.1467	217	0.60	002.06	0
002	Protein, Crude (%)	2219	24.57	0.2600	24.39	0.3022	0.1467	217	0.60	002.06	0
002	Protein, Crude (%)	0139	24.58	0.3500	24.39	0.3022	0.1467	217	0.62	002.06	0
002	Protein, Crude (%)	0425	24.58	0.1500	24.39	0.3022	0.1467	217	0.62	002.06	0
002	Protein, Crude (%)	0857	24.58	0.2300	24.39	0.3022	0.1467	217	0.62	002.06	0
002	Protein, Crude (%)	2146	24.58	0.1000	24.39	0.3022	0.1467	217	0.63	002.05	0
002	Protein, Crude (%)	2156	24.58	0.1400	24.39	0.3022	0.1467	217	0.63	002.06	0
002	Protein, Crude (%)	0309	24.59	0.1605	24.39	0.3022	0.1467	217	0.66	002.06	0
002	Protein, Crude (%)	0035	24.59	0.1600	24.39	0.3022	0.1467	217	0.67	002.06	0
002	Protein, Crude (%)	0952	24.60	0.6700	24.39	0.3022	0.1467	217	0.68	002.05	0
002	Protein, Crude (%)	0675	24.60	0.0300	24.39	0.3022	0.1467	217	0.68	002.06	0
002	Protein, Crude (%)	0610	24.60	0.2000	24.39	0.3022	0.1467	217	0.70	002.01	0
002	Protein, Crude (%)	0010	24.60	0.2000	24.39	0.3022	0.1467	217	0.70	002.06	0
002	Protein, Crude (%)	0265	24.60	0.0000	24.39	0.3022	0.1467	217	0.70	002.06	0
002	Protein, Crude (%)	0358	24.60	0.0200	24.39	0.3022	0.1467	217	0.70	002.06	0
002	Protein, Crude (%)	0520	24.60	0.0000	24.39	0.3022	0.1467	217	0.70	002.06	0
002	Protein, Crude (%)	0610	24.60	0.2000	24.39	0.3022	0.1467	217	0.70	002.06	0
002	Protein, Crude (%)	0512	24.63	0.0200	24.39	0.3022	0.1467	217	0.80	002.06	0
002	Protein, Crude (%)	2144	24.63	0.3000	24.39	0.3022	0.1467	217	0.80	002.06	0
002	Protein, Crude (%)	0004	24.64	0.1000	24.39	0.3022	0.1467	217	0.83	002.06	0
002	Protein, Crude (%)	2033	24.65	0.0500	24.39	0.3022	0.1467	217	0.85	002.06	0
002	Protein, Crude (%)	0208	24.65	0.1000	24.39	0.3022	0.1467	217	0.87	002.06	0
002	Protein, Crude (%)	0910	24.65	0.1000	24.39	0.3022	0.1467	217	0.87	002.06	0
002	Protein, Crude (%)	0160	24.65	0.5000	24.39	0.3022	0.1467	217	0.87	002.99	0
002	Protein, Crude (%)	0190	24.66	0.0500	24.39	0.3022	0.1467	217	0.88	002.06	0
002	Protein, Crude (%)	0242	24.66	0.1500	24.39	0.3022	0.1467	217	0.88	002.06	0
002	Protein, Crude (%)	0036	24.66	0.2200	24.39	0.3022	0.1467	217	0.90	002.06	0
002	Protein, Crude (%)	2128	24.67	0.0880	24.39	0.3022	0.1467	217	0.93	002.11	0
002	Protein, Crude (%)	0006	24.69	0.1100	24.39	0.3022	0.1467	217	0.98	002.06	0
002	Protein, Crude (%)	0186	24.70	0.2000	24.39	0.3022	0.1467	217	1.03	002.06	0
002	Protein, Crude (%)	0554	24.73	0.3000	24.39	0.3022	0.1467	217	1.13	002.06	0
002	Protein, Crude (%)	2207	24.73	0.4800	24.39	0.3022	0.1467	217	1.13	002.06	0
002	Protein, Crude (%)	0033	24.75	0.1000	24.39	0.3022	0.1467	217	1.20	002.06	0
002	Protein, Crude (%)	0187	24.76	0.0300	24.39	0.3022	0.1467	217	1.21	002.04	0
002	Protein, Crude (%)	0417	24.76	0.2400	24.39	0.3022	0.1467	217	1.23	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein, Crude (%)	0859	24.77	0.2670	24.39	0.3022	0.1467	217	1.27	002.06	0
002	Protein, Crude (%)	2059	24.77	0.2375	24.39	0.3022	0.1467	217	1.28	002.06	0
002	Protein, Crude (%)	0656	24.80	0.5400	24.39	0.3022	0.1467	217	1.36	002.06	0
002	Protein, Crude (%)	2089	24.80	0.1800	24.39	0.3022	0.1467	217	1.36	002.06	0
002	Protein, Crude (%)	2213	24.81	0.0900	24.39	0.3022	0.1467	217	1.38	002.06	0
002	Protein, Crude (%)	2122	24.81	0.0000	24.39	0.3022	0.1467	217	1.40	002.04	0
002	Protein, Crude (%)	0876	24.81	0.2000	24.39	0.3022	0.1467	217	1.40	002.06	0
002	Protein, Crude (%)	2076	24.81	0.2060	24.39	0.3022	0.1467	217	1.41	002.06	0
002	Protein, Crude (%)	0687	24.84	0.6800	24.39	0.3022	0.1467	217	1.50	002.06	0
002	Protein, Crude (%)	0872	24.89	0.4100	24.39	0.3022	0.1467	217	1.64	002.06	0
002	Protein, Crude (%)	2242	24.89	0.1900	24.39	0.3022	0.1467	217	1.64	002.06	0
002	Protein, Crude (%)	0407	24.89	0.0440	24.39	0.3022	0.1467	217	1.66	002.06	0
002	Protein, Crude (%)	0884	24.90	0.0000	24.39	0.3022	0.1467	217	1.69	002.06	0
002	Protein, Crude (%)	0229	24.94	0.0800	24.39	0.3022	0.1467	217	1.83	002.06	0
002	Protein, Crude (%)	2246	25.04	0.0000	24.39	0.3022	0.1467	217	2.16	002.06	0
002	Protein, Crude (%)	0609	25.12	0.1500	24.39	0.3022	0.1467	217	2.41	002.06	0
002	Protein, Crude (%)	0164	25.19	0.1200	24.39	0.3022	0.1467	217	2.65	002.01	0
002	Protein, Crude (%)	0853	25.30	0.2000	24.39	0.3022	0.1467	217	3.02	002.06	0
002	Protein, Crude (%)	2016	25.34	0.3200	24.39	0.3022	0.1467	217	3.15	002.06	0
002	Protein, Crude (%)	0034	25.38	0.3100	24.39	0.3022	0.1467	217	3.28	002.06	0
002	Protein, Crude (%)	0953	25.67	0.0100	24.39	0.3022	0.1467	217	4.23	002.06	0
002	Protein, Crude (%)	2012	25.73	0.0200	24.39	0.3022	0.1467	217	4.44	002.06	0
002	Protein, Crude (%)	0300	25.99	0.6800	24.39	0.3022	0.1467	217	5.30	002.06	0
002	Protein, Crude (%)	0918	26.67	0.3100	24.39	0.3022	0.1467	217	7.54	002.06	0
002	Protein, Crude (%)	0539	23.42	1.280	24.39	0.3022	0.1467	217	-3.20	002.06	1
002	Protein, Crude (%)	0914	23.56	0.8300	24.39	0.3022	0.1467	217	-2.76	002.01	1
002	Protein, Crude (%)	0581	24.35	0.7800	24.39	0.3022	0.1467	217	-0.13	002.06	1
002	Protein, Crude (%)	0873	19.57	0.2200	24.39	0.3022	0.1467	217	-15.95	002.06	2
003	Fat, Crude (%)	2144	4.775	0.0500	5.874	0.2447	0.1256	130	-4.49	003.14	0
003	Fat, Crude (%)	0357	4.800	0.6000	5.874	0.2447	0.1256	130	-4.39	003.14	0
003	Fat, Crude (%)	0843	5.255	0.5100	5.874	0.2447	0.1256	130	-2.53	003.14	0
003	Fat, Crude (%)	0675	5.290	0.0600	5.874	0.2447	0.1256	130	-2.38	003.09	0
003	Fat, Crude (%)	0309	5.300	0.5129	5.874	0.2447	0.1256	130	-2.34	003.00	0
003	Fat, Crude (%)	0098	5.460	0.3000	5.874	0.2447	0.1256	130	-1.69	003.13	0
003	Fat, Crude (%)	0918	5.490	0.2600	5.874	0.2447	0.1256	130	-1.57	003.06	0
003	Fat, Crude (%)	0952	5.500	0.0600	5.874	0.2447	0.1256	130	-1.53	003.10	0
003	Fat, Crude (%)	0208	5.500	0.5000	5.874	0.2447	0.1256	130	-1.53	003.13	0
003	Fat, Crude (%)	0242	5.515	0.0500	5.874	0.2447	0.1256	130	-1.47	003.10	0
003	Fat, Crude (%)	0062	5.534	0.0100	5.874	0.2447	0.1256	130	-1.39	003.10	0
003	Fat, Crude (%)	0294	5.555	0.1100	5.874	0.2447	0.1256	130	-1.30	003.06	0
003	Fat, Crude (%)	2188	5.575	0.0900	5.874	0.2447	0.1256	130	-1.22	003.10	0
003	Fat, Crude (%)	0186	5.595	0.0100	5.874	0.2447	0.1256	130	-1.14	003.14	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat, Crude (%)	2155	5.595	0.0040	5.874	0.2447	0.1256	130	-1.14	003.10	0
003	Fat, Crude (%)	0100	5.605	0.6900	5.874	0.2447	0.1256	130	-1.10	003.10	0
003	Fat, Crude (%)	2073	5.610	0.0200	5.874	0.2447	0.1256	130	-1.08	003.10	0
003	Fat, Crude (%)	0619	5.620	0.1600	5.874	0.2447	0.1256	130	-1.04	003.10	0
003	Fat, Crude (%)	0366	5.650	0.1000	5.874	0.2447	0.1256	130	-0.91	003.10	0
003	Fat, Crude (%)	2115	5.675	0.0100	5.874	0.2447	0.1256	130	-0.81	003.00	0
003	Fat, Crude (%)	0045	5.675	0.1900	5.874	0.2447	0.1256	130	-0.81	003.10	0
003	Fat, Crude (%)	0693	5.679	0.0480	5.874	0.2447	0.1256	130	-0.80	003.10	0
003	Fat, Crude (%)	2196	5.680	0.0000	5.874	0.2447	0.1256	130	-0.79	003.10	0
003	Fat, Crude (%)	0520	5.680	0.1000	5.874	0.2447	0.1256	130	-0.79	003.14	0
003	Fat, Crude (%)	2129	5.680	0.4140	5.874	0.2447	0.1256	130	-0.79	003.14	0
003	Fat, Crude (%)	2006	5.690	0.2600	5.874	0.2447	0.1256	130	-0.75	003.13	0
003	Fat, Crude (%)	0683	5.690	0.1600	5.874	0.2447	0.1256	130	-0.75	003.14	0
003	Fat, Crude (%)	0881	5.709	0.1980	5.874	0.2447	0.1256	130	-0.67	003.10	0
003	Fat, Crude (%)	2193	5.710	0.0200	5.874	0.2447	0.1256	130	-0.67	003.10	0
003	Fat, Crude (%)	2115	5.710	0.1400	5.874	0.2447	0.1256	130	-0.67	003.14	0
003	Fat, Crude (%)	0512	5.711	0.2200	5.874	0.2447	0.1256	130	-0.66	003.09	0
003	Fat, Crude (%)	0541	5.715	0.0500	5.874	0.2447	0.1256	130	-0.65	003.10	0
003	Fat, Crude (%)	0144	5.720	0.0000	5.874	0.2447	0.1256	130	-0.63	003.14	0
003	Fat, Crude (%)	0298	5.730	0.0200	5.874	0.2447	0.1256	130	-0.59	003.10	0
003	Fat, Crude (%)	0646	5.730	0.5400	5.874	0.2447	0.1256	130	-0.59	003.13	0
003	Fat, Crude (%)	0581	5.735	0.1100	5.874	0.2447	0.1256	130	-0.57	003.14	0
003	Fat, Crude (%)	0036	5.742	0.0846	5.874	0.2447	0.1256	130	-0.54	003.10	0
003	Fat, Crude (%)	0226	5.745	0.0300	5.874	0.2447	0.1256	130	-0.53	003.09	0
003	Fat, Crude (%)	0682	5.750	0.0000	5.874	0.2447	0.1256	130	-0.50	003.14	0
003	Fat, Crude (%)	0853	5.750	0.3000	5.874	0.2447	0.1256	130	-0.50	003.14	0
003	Fat, Crude (%)	0870	5.753	0.0685	5.874	0.2447	0.1256	130	-0.49	003.10	0
003	Fat, Crude (%)	0893	5.765	0.0300	5.874	0.2447	0.1256	130	-0.44	003.12	0
003	Fat, Crude (%)	0948	5.775	0.1500	5.874	0.2447	0.1256	130	-0.40	003.14	0
003	Fat, Crude (%)	0042	5.780	0.0400	5.874	0.2447	0.1256	130	-0.38	003.10	0
003	Fat, Crude (%)	0529	5.785	0.0900	5.874	0.2447	0.1256	130	-0.36	003.14	0
003	Fat, Crude (%)	2162	5.790	0.0800	5.874	0.2447	0.1256	130	-0.34	003.10	0
003	Fat, Crude (%)	0049	5.790	0.1200	5.874	0.2447	0.1256	130	-0.34	003.14	0
003	Fat, Crude (%)	0956	5.800	0.0000	5.874	0.2447	0.1256	130	-0.30	003.14	0
003	Fat, Crude (%)	0860	5.810	0.0000	5.874	0.2447	0.1256	130	-0.26	003.14	0
003	Fat, Crude (%)	0037	5.810	0.0200	5.874	0.2447	0.1256	130	-0.26	003.14	0
003	Fat, Crude (%)	2198	5.820	0.0200	5.874	0.2447	0.1256	130	-0.22	003.00	0
003	Fat, Crude (%)	0164	5.820	0.0400	5.874	0.2447	0.1256	130	-0.22	003.06	0
003	Fat, Crude (%)	0358	5.820	0.0000	5.874	0.2447	0.1256	130	-0.22	003.14	0
003	Fat, Crude (%)	0638	5.825	0.0500	5.874	0.2447	0.1256	130	-0.20	003.09	0
003	Fat, Crude (%)	0098	5.825	0.0500	5.874	0.2447	0.1256	130	-0.20	003.10	0
003	Fat, Crude (%)	2184	5.825	0.5900	5.874	0.2447	0.1256	130	-0.20	003.10	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat, Crude (%)	2246	5.830	0.0000	5.874	0.2447	0.1256	130	-0.18	003.13	0
003	Fat, Crude (%)	0425	5.835	0.0500	5.874	0.2447	0.1256	130	-0.16	003.06	0
003	Fat, Crude (%)	0354	5.835	0.0100	5.874	0.2447	0.1256	130	-0.16	003.09	0
003	Fat, Crude (%)	0726	5.835	0.0500	5.874	0.2447	0.1256	130	-0.16	003.14	0
003	Fat, Crude (%)	2146	5.835	0.3300	5.874	0.2447	0.1256	130	-0.16	003.14	0
003	Fat, Crude (%)	0027	5.836	0.1290	5.874	0.2447	0.1256	130	-0.16	003.09	0
003	Fat, Crude (%)	0878	5.845	0.0900	5.874	0.2447	0.1256	130	-0.12	003.14	0
003	Fat, Crude (%)	0674	5.850	0.0000	5.874	0.2447	0.1256	130	-0.10	003.09	0
003	Fat, Crude (%)	0689	5.850	0.1000	5.874	0.2447	0.1256	130	-0.10	003.10	0
003	Fat, Crude (%)	0278	5.850	0.1000	5.874	0.2447	0.1256	130	-0.10	003.14	0
003	Fat, Crude (%)	0407	5.856	0.1261	5.874	0.2447	0.1256	130	-0.07	003.14	0
003	Fat, Crude (%)	0083	5.860	0.0600	5.874	0.2447	0.1256	130	-0.06	003.14	0
003	Fat, Crude (%)	0504	5.865	0.2700	5.874	0.2447	0.1256	130	-0.03	003.01	0
003	Fat, Crude (%)	0948	5.865	0.0500	5.874	0.2447	0.1256	130	-0.03	003.09	0
003	Fat, Crude (%)	2018	5.865	0.0300	5.874	0.2447	0.1256	130	-0.03	003.10	0
003	Fat, Crude (%)	2190	5.870	0.2200	5.874	0.2447	0.1256	130	-0.01	003.00	0
003	Fat, Crude (%)	0108	5.870	0.0200	5.874	0.2447	0.1256	130	-0.01	003.14	0
003	Fat, Crude (%)	2009	5.872	0.0011	5.874	0.2447	0.1256	130	-0.01	003.10	0
003	Fat, Crude (%)	2109	5.880	0.0600	5.874	0.2447	0.1256	130	0.03	003.00	0
003	Fat, Crude (%)	0695	5.885	0.0300	5.874	0.2447	0.1256	130	0.05	003.10	0
003	Fat, Crude (%)	0598	5.885	0.0500	5.874	0.2447	0.1256	130	0.05	003.14	0
003	Fat, Crude (%)	2076	5.888	0.0300	5.874	0.2447	0.1256	130	0.06	003.14	0
003	Fat, Crude (%)	2192	5.905	0.0300	5.874	0.2447	0.1256	130	0.13	003.00	0
003	Fat, Crude (%)	2022	5.905	0.0500	5.874	0.2447	0.1256	130	0.13	003.10	0
003	Fat, Crude (%)	0098	5.910	0.0600	5.874	0.2447	0.1256	130	0.15	003.09	0
003	Fat, Crude (%)	0619	5.915	0.0100	5.874	0.2447	0.1256	130	0.17	003.14	0
003	Fat, Crude (%)	0187	5.920	0.0200	5.874	0.2447	0.1256	130	0.19	003.13	0
003	Fat, Crude (%)	0001	5.921	0.1050	5.874	0.2447	0.1256	130	0.19	003.14	0
003	Fat, Crude (%)	0229	5.923	0.0020	5.874	0.2447	0.1256	130	0.20	003.14	0
003	Fat, Crude (%)	0682	5.930	0.0000	5.874	0.2447	0.1256	130	0.23	003.06	0
003	Fat, Crude (%)	2207	5.930	0.0200	5.874	0.2447	0.1256	130	0.23	003.14	0
003	Fat, Crude (%)	0868	5.940	0.0200	5.874	0.2447	0.1256	130	0.27	003.14	0
003	Fat, Crude (%)	0142	5.955	0.0100	5.874	0.2447	0.1256	130	0.33	003.00	0
003	Fat, Crude (%)	0353	5.955	0.0100	5.874	0.2447	0.1256	130	0.33	003.09	0
003	Fat, Crude (%)	0229	5.958	0.0100	5.874	0.2447	0.1256	130	0.35	003.06	0
003	Fat, Crude (%)	0590	5.965	0.1900	5.874	0.2447	0.1256	130	0.37	003.09	0
003	Fat, Crude (%)	0897	5.970	0.0400	5.874	0.2447	0.1256	130	0.39	003.10	0
003	Fat, Crude (%)	0066	5.985	0.2300	5.874	0.2447	0.1256	130	0.46	003.14	0
003	Fat, Crude (%)	0849	5.992	0.0970	5.874	0.2447	0.1256	130	0.48	003.06	0
003	Fat, Crude (%)	0205	6.001	0.2030	5.874	0.2447	0.1256	130	0.52	003.13	0
003	Fat, Crude (%)	0644	6.014	0.1725	5.874	0.2447	0.1256	130	0.57	003.06	0
003	Fat, Crude (%)	0171	6.015	0.0500	5.874	0.2447	0.1256	130	0.58	003.12	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat, Crude (%)	0345	6.030	0.0400	5.874	0.2447	0.1256	130	0.64	003.00	0
003	Fat, Crude (%)	0190	6.030	0.0800	5.874	0.2447	0.1256	130	0.64	003.14	0
003	Fat, Crude (%)	0910	6.035	0.1100	5.874	0.2447	0.1256	130	0.66	003.06	0
003	Fat, Crude (%)	0613	6.035	0.7500	5.874	0.2447	0.1256	130	0.66	003.14	0
003	Fat, Crude (%)	0689	6.050	0.1000	5.874	0.2447	0.1256	130	0.72	003.06	0
003	Fat, Crude (%)	0955	6.050	0.1800	5.874	0.2447	0.1256	130	0.72	003.14	0
003	Fat, Crude (%)	0939	6.055	0.4100	5.874	0.2447	0.1256	130	0.74	003.14	0
003	Fat, Crude (%)	0265	6.085	0.1300	5.874	0.2447	0.1256	130	0.86	003.14	0
003	Fat, Crude (%)	0202	6.086	0.1390	5.874	0.2447	0.1256	130	0.87	003.14	0
003	Fat, Crude (%)	0089	6.130	0.0000	5.874	0.2447	0.1256	130	1.05	003.14	0
003	Fat, Crude (%)	0884	6.150	0.3000	5.874	0.2447	0.1256	130	1.13	003.06	0
003	Fat, Crude (%)	2012	6.160	0.2800	5.874	0.2447	0.1256	130	1.17	003.09	0
003	Fat, Crude (%)	0968	6.173	0.0870	5.874	0.2447	0.1256	130	1.22	003.14	0
003	Fat, Crude (%)	0885	6.189	0.0282	5.874	0.2447	0.1256	130	1.29	003.00	0
003	Fat, Crude (%)	0676	6.235	0.1900	5.874	0.2447	0.1256	130	1.48	003.10	0
003	Fat, Crude (%)	0609	6.500	0.0000	5.874	0.2447	0.1256	130	2.56	003.10	0
003	Fat, Crude (%)	2126	6.545	0.0700	5.874	0.2447	0.1256	130	2.74	003.00	0
003	Fat, Crude (%)	0546	6.575	0.2900	5.874	0.2447	0.1256	130	2.87	003.99	0
003	Fat, Crude (%)	2222	6.580	0.0600	5.874	0.2447	0.1256	130	2.89	003.10	0
003	Fat, Crude (%)	0889	7.310	0.0200	5.874	0.2447	0.1256	130	5.87	003.11	0
003	Fat, Crude (%)	0613	7.860	0.0800	5.874	0.2447	0.1256	130	8.12	003.11	0
003	Fat, Crude (%)	0011	7.950	0.3000	5.874	0.2447	0.1256	130	8.49	003.11	0
003	Fat, Crude (%)	0169	8.010	0.0000	5.874	0.2447	0.1256	130	8.73	003.06	0
003	Fat, Crude (%)	2144	8.420	0.0600	5.874	0.2447	0.1256	130	10.41	003.99	0
003	Fat, Crude (%)	0034	8.505	0.0300	5.874	0.2447	0.1256	130	10.75	003.13	0
003	Fat, Crude (%)	0662	8.632	0.0042	5.874	0.2447	0.1256	130	11.27	003.06	0
003	Fat, Crude (%)	0941	8.840	0.1200	5.874	0.2447	0.1256	130	12.12	003.99	0
003	Fat, Crude (%)	0032	9.150	0.3000	5.874	0.2447	0.1256	130	13.39	003.00	0
003	Fat, Crude (%)	0405	9.150	0.0000	5.874	0.2447	0.1256	130	13.39	003.11	0
003	Fat, Crude (%)	2128	9.511	0.0049	5.874	0.2447	0.1256	130	14.86	003.11	0
003	Fat, Crude (%)	0656	9.710	0.4800	5.874	0.2447	0.1256	130	15.68	003.06	0
003	Fat, Crude (%)	0852	9.885	0.2100	5.874	0.2447	0.1256	130	16.39	003.11	0
003	Fat, Crude (%)	0889	4.215	2.490	5.874	0.2447	0.1256	130	-6.78	003.10	1
003	Fat, Crude (%)	0687	6.500	1.000	5.874	0.2447	0.1256	130	2.56	003.06	1
004	Fiber, Crude (%)	0202	1.930	0.0060	3.511	0.4446	0.1575	132	-3.56	004.07	0
004	Fiber, Crude (%)	2089	2.010	0.0600	3.511	0.4446	0.1575	132	-3.38	004.03	0
004	Fiber, Crude (%)	0405	2.240	0.0000	3.511	0.4446	0.1575	132	-2.86	004.11	0
004	Fiber, Crude (%)	0598	2.385	0.0300	3.511	0.4446	0.1575	132	-2.53	004.07	0
004	Fiber, Crude (%)	0042	2.610	0.7600	3.511	0.4446	0.1575	132	-2.03	004.07	0
004	Fiber, Crude (%)	2146	2.775	0.0500	3.511	0.4446	0.1575	132	-1.66	004.99	0
004	Fiber, Crude (%)	0004	2.865	0.2900	3.511	0.4446	0.1575	132	-1.45	004.07	0
004	Fiber, Crude (%)	0619	2.895	0.0700	3.511	0.4446	0.1575	132	-1.39	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber, Crude (%)	0294	2.900	0.0000	3.511	0.4446	0.1575	132	-1.38	004.07	0
004	Fiber, Crude (%)	0968	2.910	0.3000	3.511	0.4446	0.1575	132	-1.35	004.07	0
004	Fiber, Crude (%)	0510	2.950	0.1000	3.511	0.4446	0.1575	132	-1.26	004.07	0
004	Fiber, Crude (%)	0644	2.969	0.3180	3.511	0.4446	0.1575	132	-1.22	004.07	0
004	Fiber, Crude (%)	0921	2.990	0.3200	3.511	0.4446	0.1575	132	-1.17	004.03	0
004	Fiber, Crude (%)	0682	3.000	0.0000	3.511	0.4446	0.1575	132	-1.15	004.07	0
004	Fiber, Crude (%)	0939	3.000	0.3800	3.511	0.4446	0.1575	132	-1.15	004.07	0
004	Fiber, Crude (%)	0003	3.040	0.3400	3.511	0.4446	0.1575	132	-1.06	004.07	0
004	Fiber, Crude (%)	2073	3.050	0.0600	3.511	0.4446	0.1575	132	-1.04	004.00	0
004	Fiber, Crude (%)	0520	3.050	0.1000	3.511	0.4446	0.1575	132	-1.04	004.07	0
004	Fiber, Crude (%)	2033	3.050	0.1000	3.511	0.4446	0.1575	132	-1.04	004.07	0
004	Fiber, Crude (%)	0407	3.056	0.0298	3.511	0.4446	0.1575	132	-1.02	004.07	0
004	Fiber, Crude (%)	0708	3.070	0.0600	3.511	0.4446	0.1575	132	-0.99	004.07	0
004	Fiber, Crude (%)	0098	3.085	0.3300	3.511	0.4446	0.1575	132	-0.96	004.07	0
004	Fiber, Crude (%)	0169	3.095	0.2100	3.511	0.4446	0.1575	132	-0.94	004.00	0
004	Fiber, Crude (%)	0100	3.100	0.2200	3.511	0.4446	0.1575	132	-0.93	004.07	0
004	Fiber, Crude (%)	2009	3.145	0.0358	3.511	0.4446	0.1575	132	-0.82	004.07	0
004	Fiber, Crude (%)	2146	3.185	0.0900	3.511	0.4446	0.1575	132	-0.73	004.07	0
004	Fiber, Crude (%)	0870	3.194	0.1567	3.511	0.4446	0.1575	132	-0.71	004.07	0
004	Fiber, Crude (%)	0653	3.200	0.1200	3.511	0.4446	0.1575	132	-0.70	004.07	0
004	Fiber, Crude (%)	0009	3.215	0.0700	3.511	0.4446	0.1575	132	-0.67	004.07	0
004	Fiber, Crude (%)	0726	3.230	0.0400	3.511	0.4446	0.1575	132	-0.63	004.07	0
004	Fiber, Crude (%)	0675	3.240	0.3000	3.511	0.4446	0.1575	132	-0.61	004.07	0
004	Fiber, Crude (%)	2109	3.245	0.0100	3.511	0.4446	0.1575	132	-0.60	004.07	0
004	Fiber, Crude (%)	2058	3.250	0.1000	3.511	0.4446	0.1575	132	-0.59	004.03	0
004	Fiber, Crude (%)	0208	3.265	0.1100	3.511	0.4446	0.1575	132	-0.55	004.00	0
004	Fiber, Crude (%)	0032	3.270	0.1400	3.511	0.4446	0.1575	132	-0.54	004.07	0
004	Fiber, Crude (%)	0504	3.275	0.2900	3.511	0.4446	0.1575	132	-0.53	004.00	0
004	Fiber, Crude (%)	0229	3.275	0.0300	3.511	0.4446	0.1575	132	-0.53	004.07	0
004	Fiber, Crude (%)	2053	3.280	0.2400	3.511	0.4446	0.1575	132	-0.52	004.07	0
004	Fiber, Crude (%)	0683	3.290	0.1200	3.511	0.4446	0.1575	132	-0.50	004.07	0
004	Fiber, Crude (%)	0171	3.295	0.2700	3.511	0.4446	0.1575	132	-0.49	004.00	0
004	Fiber, Crude (%)	0298	3.300	0.2000	3.511	0.4446	0.1575	132	-0.48	004.00	0
004	Fiber, Crude (%)	0278	3.300	0.0000	3.511	0.4446	0.1575	132	-0.48	004.07	0
004	Fiber, Crude (%)	0956	3.300	0.4000	3.511	0.4446	0.1575	132	-0.48	004.07	0
004	Fiber, Crude (%)	2190	3.315	0.1500	3.511	0.4446	0.1575	132	-0.44	004.07	0
004	Fiber, Crude (%)	0881	3.327	0.0640	3.511	0.4446	0.1575	132	-0.41	004.06	0
004	Fiber, Crude (%)	0083	3.330	0.0200	3.511	0.4446	0.1575	132	-0.41	004.07	0
004	Fiber, Crude (%)	2165	3.340	0.0000	3.511	0.4446	0.1575	132	-0.39	004.00	0
004	Fiber, Crude (%)	0098	3.340	0.1400	3.511	0.4446	0.1575	132	-0.39	004.06	0
004	Fiber, Crude (%)	2006	3.345	0.0300	3.511	0.4446	0.1575	132	-0.37	004.06	0
004	Fiber, Crude (%)	0590	3.350	0.0600	3.511	0.4446	0.1575	132	-0.36	004.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber, Crude (%)	0910	3.350	0.1000	3.511	0.4446	0.1575	132	-0.36	004.07	0
004	Fiber, Crude (%)	0695	3.350	0.2800	3.511	0.4446	0.1575	132	-0.36	004.00	0
004	Fiber, Crude (%)	0144	3.370	0.2000	3.511	0.4446	0.1575	132	-0.32	004.07	0
004	Fiber, Crude (%)	2190	3.385	0.0700	3.511	0.4446	0.1575	132	-0.28	004.03	0
004	Fiber, Crude (%)	0512	3.401	0.1130	3.511	0.4446	0.1575	132	-0.25	004.06	0
004	Fiber, Crude (%)	0353	3.410	0.0200	3.511	0.4446	0.1575	132	-0.23	004.03	0
004	Fiber, Crude (%)	0613	3.410	0.0400	3.511	0.4446	0.1575	132	-0.23	004.11	0
004	Fiber, Crude (%)	2128	3.425	0.0479	3.511	0.4446	0.1575	132	-0.20	004.11	0
004	Fiber, Crude (%)	0345	3.430	0.1400	3.511	0.4446	0.1575	132	-0.18	004.00	0
004	Fiber, Crude (%)	0918	3.440	0.3000	3.511	0.4446	0.1575	132	-0.16	004.06	0
004	Fiber, Crude (%)	0190	3.445	0.0900	3.511	0.4446	0.1575	132	-0.15	004.07	0
004	Fiber, Crude (%)	0164	3.450	0.1000	3.511	0.4446	0.1575	132	-0.14	004.00	0
004	Fiber, Crude (%)	0425	3.450	0.1000	3.511	0.4446	0.1575	132	-0.14	004.00	0
004	Fiber, Crude (%)	0689	3.450	0.1000	3.511	0.4446	0.1575	132	-0.14	004.06	0
004	Fiber, Crude (%)	0074	3.450	0.2600	3.511	0.4446	0.1575	132	-0.14	004.07	0
004	Fiber, Crude (%)	2144	3.450	0.0600	3.511	0.4446	0.1575	132	-0.14	004.07	0
004	Fiber, Crude (%)	2181	3.450	0.1400	3.511	0.4446	0.1575	132	-0.14	004.07	0
004	Fiber, Crude (%)	2076	3.489	0.4756	3.511	0.4446	0.1575	132	-0.05	004.00	0
004	Fiber, Crude (%)	0205	3.490	0.0600	3.511	0.4446	0.1575	132	-0.05	004.06	0
004	Fiber, Crude (%)	0019	3.500	0.3600	3.511	0.4446	0.1575	132	-0.03	004.07	0
004	Fiber, Crude (%)	0354	3.500	0.0400	3.511	0.4446	0.1575	132	-0.03	004.07	0
004	Fiber, Crude (%)	0309	3.500	0.0057	3.511	0.4446	0.1575	132	-0.03	004.00	0
004	Fiber, Crude (%)	0675	3.510	0.5600	3.511	0.4446	0.1575	132	0.00	004.06	0
004	Fiber, Crude (%)	0849	3.514	0.0480	3.511	0.4446	0.1575	132	0.01	004.07	0
004	Fiber, Crude (%)	2124	3.520	0.0200	3.511	0.4446	0.1575	132	0.02	004.07	0
004	Fiber, Crude (%)	0848	3.520	0.0000	3.511	0.4446	0.1575	132	0.02	004.07	0
004	Fiber, Crude (%)	0263	3.543	0.0397	3.511	0.4446	0.1575	132	0.07	004.07	0
004	Fiber, Crude (%)	0028	3.550	0.1000	3.511	0.4446	0.1575	132	0.09	004.07	0
004	Fiber, Crude (%)	0581	3.550	0.3000	3.511	0.4446	0.1575	132	0.09	004.07	0
004	Fiber, Crude (%)	0968	3.558	0.0070	3.511	0.4446	0.1575	132	0.10	004.06	0
004	Fiber, Crude (%)	2022	3.585	0.0100	3.511	0.4446	0.1575	132	0.17	004.06	0
004	Fiber, Crude (%)	0242	3.590	0.5200	3.511	0.4446	0.1575	132	0.18	004.07	0
004	Fiber, Crude (%)	0300	3.590	0.1400	3.511	0.4446	0.1575	132	0.18	004.07	0
004	Fiber, Crude (%)	0265	3.595	0.7100	3.511	0.4446	0.1575	132	0.19	004.07	0
004	Fiber, Crude (%)	0878	3.600	0.0400	3.511	0.4446	0.1575	132	0.20	004.06	0
004	Fiber, Crude (%)	0610	3.600	0.2000	3.511	0.4446	0.1575	132	0.20	004.99	0
004	Fiber, Crude (%)	0034	3.605	0.2100	3.511	0.4446	0.1575	132	0.21	004.07	0
004	Fiber, Crude (%)	0354	3.610	0.0000	3.511	0.4446	0.1575	132	0.22	004.00	0
004	Fiber, Crude (%)	0948	3.625	0.0500	3.511	0.4446	0.1575	132	0.26	004.07	0
004	Fiber, Crude (%)	0876	3.673	0.5704	3.511	0.4446	0.1575	132	0.36	004.07	0
004	Fiber, Crude (%)	0626	3.685	0.1700	3.511	0.4446	0.1575	132	0.39	004.07	0
004	Fiber, Crude (%)	2207	3.690	0.2000	3.511	0.4446	0.1575	132	0.40	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber, Crude (%)	0897	3.690	0.0200	3.511	0.4446	0.1575	132	0.40	004.06	0
004	Fiber, Crude (%)	0638	3.700	0.2000	3.511	0.4446	0.1575	132	0.42	004.06	0
004	Fiber, Crude (%)	2115	3.710	0.0200	3.511	0.4446	0.1575	132	0.45	004.06	0
004	Fiber, Crude (%)	0529	3.710	0.0200	3.511	0.4446	0.1575	132	0.45	004.07	0
004	Fiber, Crude (%)	0948	3.745	0.0900	3.511	0.4446	0.1575	132	0.53	004.06	0
004	Fiber, Crude (%)	0885	3.751	0.0612	3.511	0.4446	0.1575	132	0.54	004.06	0
004	Fiber, Crude (%)	0089	3.770	0.1200	3.511	0.4446	0.1575	132	0.58	004.07	0
004	Fiber, Crude (%)	0096	3.775	0.0500	3.511	0.4446	0.1575	132	0.59	004.07	0
004	Fiber, Crude (%)	2196	3.790	0.0000	3.511	0.4446	0.1575	132	0.63	004.07	0
004	Fiber, Crude (%)	0027	3.830	0.0400	3.511	0.4446	0.1575	132	0.72	004.06	0
004	Fiber, Crude (%)	0941	3.840	0.1000	3.511	0.4446	0.1575	132	0.74	004.06	0
004	Fiber, Crude (%)	2188	3.845	0.0700	3.511	0.4446	0.1575	132	0.75	004.07	0
004	Fiber, Crude (%)	0868	3.855	0.0300	3.511	0.4446	0.1575	132	0.77	004.06	0
004	Fiber, Crude (%)	0066	3.870	0.3600	3.511	0.4446	0.1575	132	0.81	004.07	0
004	Fiber, Crude (%)	0952	3.880	0.2600	3.511	0.4446	0.1575	132	0.83	004.06	0
004	Fiber, Crude (%)	2192	3.925	0.2300	3.511	0.4446	0.1575	132	0.93	004.03	0
004	Fiber, Crude (%)	0955	3.935	0.4100	3.511	0.4446	0.1575	132	0.95	004.07	0
004	Fiber, Crude (%)	0596	3.965	0.6300	3.511	0.4446	0.1575	132	1.02	004.07	0
004	Fiber, Crude (%)	0505	3.985	0.2500	3.511	0.4446	0.1575	132	1.07	004.07	0
004	Fiber, Crude (%)	2144	3.995	0.1100	3.511	0.4446	0.1575	132	1.09	004.99	0
004	Fiber, Crude (%)	0017	4.100	0.2000	3.511	0.4446	0.1575	132	1.32	004.00	0
004	Fiber, Crude (%)	2259	4.120	0.1000	3.511	0.4446	0.1575	132	1.37	004.07	0
004	Fiber, Crude (%)	0860	4.135	0.0000	3.511	0.4446	0.1575	132	1.40	004.06	0
004	Fiber, Crude (%)	0693	4.283	0.6040	3.511	0.4446	0.1575	132	1.74	004.07	0
004	Fiber, Crude (%)	0609	4.310	0.0200	3.511	0.4446	0.1575	132	1.80	004.06	0
004	Fiber, Crude (%)	0613	4.335	0.5700	3.511	0.4446	0.1575	132	1.85	004.07	0
004	Fiber, Crude (%)	2198	4.385	0.0100	3.511	0.4446	0.1575	132	1.97	004.06	0
004	Fiber, Crude (%)	0511	4.400	0.2000	3.511	0.4446	0.1575	132	2.00	004.00	0
004	Fiber, Crude (%)	2012	4.450	0.1000	3.511	0.4446	0.1575	132	2.11	004.07	0
004	Fiber, Crude (%)	0884	4.500	0.2000	3.511	0.4446	0.1575	132	2.22	004.00	0
004	Fiber, Crude (%)	0845	4.525	0.0500	3.511	0.4446	0.1575	132	2.28	004.06	0
004	Fiber, Crude (%)	0656	4.675	0.3300	3.511	0.4446	0.1575	132	2.62	004.07	0
004	Fiber, Crude (%)	0646	4.745	0.1100	3.511	0.4446	0.1575	132	2.77	004.07	0
004	Fiber, Crude (%)	2193	4.975	0.0700	3.511	0.4446	0.1575	132	3.29	004.07	0
004	Fiber, Crude (%)	0676	5.025	0.3700	3.511	0.4446	0.1575	132	3.40	004.06	0
004	Fiber, Crude (%)	0674	5.320	0.0000	3.511	0.4446	0.1575	132	4.07	004.06	0
004	Fiber, Crude (%)	0889	5.325	0.0300	3.511	0.4446	0.1575	132	4.08	004.11	0
004	Fiber, Crude (%)	2246	5.345	0.0500	3.511	0.4446	0.1575	132	4.12	004.03	0
004	Fiber, Crude (%)	0226	5.960	0.3200	3.511	0.4446	0.1575	132	5.51	004.00	0
004	Fiber, Crude (%)	0893	7.395	0.0700	3.511	0.4446	0.1575	132	8.74	004.06	0
004	Fiber, Crude (%)	0366	2.600	1.400	3.511	0.4446	0.1575	132	-2.05	004.01	1
004	Fiber, Crude (%)	0045	4.415	0.8700	3.511	0.4446	0.1575	132	2.03	004.07	1

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0873	5.840	0.0400	7.655	0.1835	0.0703	159	-9.89	005.00	0
005	Ash (%)	2081	6.805	0.1700	7.655	0.1835	0.0703	159	-4.63	005.99	0
005	Ash (%)	0889	6.865	0.0100	7.655	0.1835	0.0703	159	-4.31	005.11	0
005	Ash (%)	2128	6.885	0.3418	7.655	0.1835	0.0703	159	-4.20	005.11	0
005	Ash (%)	2169	7.260	0.3200	7.655	0.1835	0.0703	159	-2.15	005.99	0
005	Ash (%)	0874	7.265	0.0700	7.655	0.1835	0.0703	159	-2.12	005.00	0
005	Ash (%)	0853	7.300	0.2000	7.655	0.1835	0.0703	159	-1.93	005.00	0
005	Ash (%)	0674	7.320	0.0000	7.655	0.1835	0.0703	159	-1.83	005.00	0
005	Ash (%)	2184	7.325	0.0300	7.655	0.1835	0.0703	159	-1.80	005.00	0
005	Ash (%)	2073	7.340	0.0200	7.655	0.1835	0.0703	159	-1.72	005.00	0
005	Ash (%)	0539	7.345	0.1500	7.655	0.1835	0.0703	159	-1.69	005.00	0
005	Ash (%)	0953	7.370	0.1200	7.655	0.1835	0.0703	159	-1.55	005.00	0
005	Ash (%)	0662	7.374	0.1859	7.655	0.1835	0.0703	159	-1.53	005.99	0
005	Ash (%)	0004	7.390	0.0400	7.655	0.1835	0.0703	159	-1.44	005.00	0
005	Ash (%)	0651	7.399	0.0320	7.655	0.1835	0.0703	159	-1.39	005.00	0
005	Ash (%)	2033	7.400	0.0000	7.655	0.1835	0.0703	159	-1.39	005.00	0
005	Ash (%)	0957	7.405	0.2100	7.655	0.1835	0.0703	159	-1.36	005.00	0
005	Ash (%)	0169	7.410	0.0800	7.655	0.1835	0.0703	159	-1.33	005.00	0
005	Ash (%)	0027	7.413	0.0210	7.655	0.1835	0.0703	159	-1.32	005.00	0
005	Ash (%)	0417	7.420	0.0400	7.655	0.1835	0.0703	159	-1.28	005.00	0
005	Ash (%)	2076	7.421	0.0225	7.655	0.1835	0.0703	159	-1.27	005.00	0
005	Ash (%)	0589	7.440	0.0400	7.655	0.1835	0.0703	159	-1.17	005.00	0
005	Ash (%)	0682	7.440	0.0000	7.655	0.1835	0.0703	159	-1.17	005.00	0
005	Ash (%)	0726	7.440	0.0000	7.655	0.1835	0.0703	159	-1.17	005.05	0
005	Ash (%)	0278	7.450	0.1000	7.655	0.1835	0.0703	159	-1.12	005.00	0
005	Ash (%)	0956	7.450	0.1000	7.655	0.1835	0.0703	159	-1.12	005.00	0
005	Ash (%)	0229	7.454	0.0550	7.655	0.1835	0.0703	159	-1.10	005.00	0
005	Ash (%)	0208	7.455	0.0300	7.655	0.1835	0.0703	159	-1.09	005.00	0
005	Ash (%)	0520	7.455	0.0700	7.655	0.1835	0.0703	159	-1.09	005.00	0
005	Ash (%)	0049	7.460	0.0200	7.655	0.1835	0.0703	159	-1.06	005.00	0
005	Ash (%)	0643	7.460	0.1000	7.655	0.1835	0.0703	159	-1.06	005.00	0
005	Ash (%)	0638	7.480	0.0400	7.655	0.1835	0.0703	159	-0.95	005.05	0
005	Ash (%)	0881	7.483	0.0270	7.655	0.1835	0.0703	159	-0.94	005.00	0
005	Ash (%)	0650	7.495	0.1100	7.655	0.1835	0.0703	159	-0.87	005.00	0
005	Ash (%)	0047	7.500	0.0000	7.655	0.1835	0.0703	159	-0.84	005.00	0
005	Ash (%)	0541	7.500	0.0600	7.655	0.1835	0.0703	159	-0.84	005.00	0
005	Ash (%)	0019	7.505	0.0100	7.655	0.1835	0.0703	159	-0.82	005.00	0
005	Ash (%)	0510	7.505	0.0100	7.655	0.1835	0.0703	159	-0.82	005.00	0
005	Ash (%)	0876	7.508	0.0650	7.655	0.1835	0.0703	159	-0.80	005.00	0
005	Ash (%)	0066	7.510	0.0800	7.655	0.1835	0.0703	159	-0.79	005.05	0
005	Ash (%)	0964	7.518	0.0830	7.655	0.1835	0.0703	159	-0.75	005.00	0
005	Ash (%)	0142	7.520	0.0000	7.655	0.1835	0.0703	159	-0.73	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0918	7.520	0.0600	7.655	0.1835	0.0703	159	-0.73	005.00	0
005	Ash (%)	0354	7.530	0.0200	7.655	0.1835	0.0703	159	-0.68	005.00	0
005	Ash (%)	0034	7.531	0.0710	7.655	0.1835	0.0703	159	-0.68	005.00	0
005	Ash (%)	0860	7.531	0.0000	7.655	0.1835	0.0703	159	-0.68	005.00	0
005	Ash (%)	0511	7.535	0.0100	7.655	0.1835	0.0703	159	-0.65	005.00	0
005	Ash (%)	0598	7.540	0.1000	7.655	0.1835	0.0703	159	-0.63	005.00	0
005	Ash (%)	0051	7.550	0.1000	7.655	0.1835	0.0703	159	-0.57	005.00	0
005	Ash (%)	0529	7.550	0.0000	7.655	0.1835	0.0703	159	-0.57	005.00	0
005	Ash (%)	0693	7.569	0.0380	7.655	0.1835	0.0703	159	-0.47	005.00	0
005	Ash (%)	0045	7.570	0.1000	7.655	0.1835	0.0703	159	-0.46	005.00	0
005	Ash (%)	2022	7.585	0.1100	7.655	0.1835	0.0703	159	-0.38	005.00	0
005	Ash (%)	0003	7.585	0.0700	7.655	0.1835	0.0703	159	-0.38	005.05	0
005	Ash (%)	0358	7.595	0.0500	7.655	0.1835	0.0703	159	-0.33	005.00	0
005	Ash (%)	0407	7.599	0.1853	7.655	0.1835	0.0703	159	-0.31	005.00	0
005	Ash (%)	0512	7.603	0.1190	7.655	0.1835	0.0703	159	-0.29	005.00	0
005	Ash (%)	0298	7.610	0.0200	7.655	0.1835	0.0703	159	-0.24	005.00	0
005	Ash (%)	0870	7.617	0.1667	7.655	0.1835	0.0703	159	-0.20	005.00	0
005	Ash (%)	2188	7.620	0.0200	7.655	0.1835	0.0703	159	-0.19	005.00	0
005	Ash (%)	2220	7.625	0.0300	7.655	0.1835	0.0703	159	-0.16	005.00	0
005	Ash (%)	0001	7.627	0.0901	7.655	0.1835	0.0703	159	-0.15	005.00	0
005	Ash (%)	0108	7.630	0.1400	7.655	0.1835	0.0703	159	-0.14	005.00	0
005	Ash (%)	0845	7.630	0.1200	7.655	0.1835	0.0703	159	-0.14	005.05	0
005	Ash (%)	0893	7.635	0.0100	7.655	0.1835	0.0703	159	-0.11	005.05	0
005	Ash (%)	0646	7.645	0.1700	7.655	0.1835	0.0703	159	-0.05	005.00	0
005	Ash (%)	2156	7.645	0.0700	7.655	0.1835	0.0703	159	-0.05	005.00	0
005	Ash (%)	0033	7.650	0.0000	7.655	0.1835	0.0703	159	-0.03	005.00	0
005	Ash (%)	2016	7.650	0.0400	7.655	0.1835	0.0703	159	-0.03	005.00	0
005	Ash (%)	2162	7.650	0.1800	7.655	0.1835	0.0703	159	-0.03	005.00	0
005	Ash (%)	0958	7.655	0.0100	7.655	0.1835	0.0703	159	0.00	005.00	0
005	Ash (%)	0613	7.655	0.0900	7.655	0.1835	0.0703	159	0.00	005.11	0
005	Ash (%)	2006	7.660	0.0200	7.655	0.1835	0.0703	159	0.03	005.00	0
005	Ash (%)	2242	7.665	0.0700	7.655	0.1835	0.0703	159	0.06	005.00	0
005	Ash (%)	0083	7.670	0.0600	7.655	0.1835	0.0703	159	0.08	005.00	0
005	Ash (%)	2053	7.670	0.0400	7.655	0.1835	0.0703	159	0.08	005.00	0
005	Ash (%)	0546	7.675	0.0500	7.655	0.1835	0.0703	159	0.11	005.99	0
005	Ash (%)	2259	7.685	0.1500	7.655	0.1835	0.0703	159	0.16	005.00	0
005	Ash (%)	2264	7.688	0.0427	7.655	0.1835	0.0703	159	0.18	005.05	0
005	Ash (%)	2150	7.695	0.0700	7.655	0.1835	0.0703	159	0.22	005.00	0
005	Ash (%)	2196	7.698	0.0000	7.655	0.1835	0.0703	159	0.24	005.00	0
005	Ash (%)	0889	7.700	0.1000	7.655	0.1835	0.0703	159	0.25	005.05	0
005	Ash (%)	0098	7.705	0.0500	7.655	0.1835	0.0703	159	0.27	005.00	0
005	Ash (%)	0035	7.710	0.0600	7.655	0.1835	0.0703	159	0.30	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0187	7.710	0.0200	7.655	0.1835	0.0703	159	0.30	005.00	0
005	Ash (%)	0653	7.710	0.1200	7.655	0.1835	0.0703	159	0.30	005.00	0
005	Ash (%)	0139	7.715	0.0300	7.655	0.1835	0.0703	159	0.33	005.00	0
005	Ash (%)	0164	7.715	0.2700	7.655	0.1835	0.0703	159	0.33	005.00	0
005	Ash (%)	2146	7.715	0.1300	7.655	0.1835	0.0703	159	0.33	005.05	0
005	Ash (%)	0968	7.720	0.0260	7.655	0.1835	0.0703	159	0.36	005.00	0
005	Ash (%)	2198	7.720	0.0400	7.655	0.1835	0.0703	159	0.36	005.05	0
005	Ash (%)	0263	7.724	0.0030	7.655	0.1835	0.0703	159	0.38	005.05	0
005	Ash (%)	2058	7.725	0.0300	7.655	0.1835	0.0703	159	0.38	005.00	0
005	Ash (%)	0868	7.725	0.0100	7.655	0.1835	0.0703	159	0.38	005.05	0
005	Ash (%)	0345	7.730	0.2400	7.655	0.1835	0.0703	159	0.41	005.00	0
005	Ash (%)	0656	7.730	0.0400	7.655	0.1835	0.0703	159	0.41	005.00	0
005	Ash (%)	0675	7.730	0.0600	7.655	0.1835	0.0703	159	0.41	005.00	0
005	Ash (%)	2252	7.730	0.0200	7.655	0.1835	0.0703	159	0.41	005.00	0
005	Ash (%)	0596	7.735	0.1100	7.655	0.1835	0.0703	159	0.44	005.00	0
005	Ash (%)	2207	7.735	0.0100	7.655	0.1835	0.0703	159	0.44	005.00	0
005	Ash (%)	2219	7.745	0.0100	7.655	0.1835	0.0703	159	0.49	005.00	0
005	Ash (%)	0644	7.745	0.0300	7.655	0.1835	0.0703	159	0.49	005.02	0
005	Ash (%)	2144	7.745	0.0100	7.655	0.1835	0.0703	159	0.49	005.05	0
005	Ash (%)	0676	7.745	0.0300	7.655	0.1835	0.0703	159	0.49	005.99	0
005	Ash (%)	0265	7.750	0.1000	7.655	0.1835	0.0703	159	0.52	005.00	0
005	Ash (%)	0366	7.750	0.1000	7.655	0.1835	0.0703	159	0.52	005.00	0
005	Ash (%)	0581	7.750	0.1200	7.655	0.1835	0.0703	159	0.52	005.00	0
005	Ash (%)	0144	7.755	0.0100	7.655	0.1835	0.0703	159	0.55	005.00	0
005	Ash (%)	0505	7.755	0.0100	7.655	0.1835	0.0703	159	0.55	005.00	0
005	Ash (%)	0609	7.765	0.0900	7.655	0.1835	0.0703	159	0.60	005.00	0
005	Ash (%)	0848	7.765	0.0100	7.655	0.1835	0.0703	159	0.60	005.00	0
005	Ash (%)	0171	7.765	0.1100	7.655	0.1835	0.0703	159	0.60	005.00	0
005	Ash (%)	0914	7.765	0.0700	7.655	0.1835	0.0703	159	0.60	005.00	0
005	Ash (%)	0948	7.765	0.0300	7.655	0.1835	0.0703	159	0.60	005.00	0
005	Ash (%)	0878	7.770	0.0800	7.655	0.1835	0.0703	159	0.63	005.05	0
005	Ash (%)	2109	7.770	0.1200	7.655	0.1835	0.0703	159	0.63	005.05	0
005	Ash (%)	2246	7.775	0.0500	7.655	0.1835	0.0703	159	0.65	005.00	0
005	Ash (%)	2009	7.778	0.0120	7.655	0.1835	0.0703	159	0.67	005.05	0
005	Ash (%)	2115	7.780	0.0200	7.655	0.1835	0.0703	159	0.68	005.00	0
005	Ash (%)	0885	7.783	0.0477	7.655	0.1835	0.0703	159	0.70	005.05	0
005	Ash (%)	2012	7.785	0.0300	7.655	0.1835	0.0703	159	0.71	005.05	0
005	Ash (%)	0202	7.789	0.0750	7.655	0.1835	0.0703	159	0.73	005.99	0
005	Ash (%)	0242	7.790	0.0600	7.655	0.1835	0.0703	159	0.74	005.00	0
005	Ash (%)	0689	7.790	0.0200	7.655	0.1835	0.0703	159	0.74	005.05	0
005	Ash (%)	0897	7.790	0.0200	7.655	0.1835	0.0703	159	0.74	005.05	0
005	Ash (%)	2246	7.790	0.0600	7.655	0.1835	0.0703	159	0.74	005.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0425	7.795	0.0300	7.655	0.1835	0.0703	159	0.76	005.00	0
005	Ash (%)	0294	7.800	0.0000	7.655	0.1835	0.0703	159	0.79	005.05	0
005	Ash (%)	0017	7.800	0.2000	7.655	0.1835	0.0703	159	0.79	005.00	0
005	Ash (%)	0884	7.800	0.2000	7.655	0.1835	0.0703	159	0.79	005.99	0
005	Ash (%)	0849	7.801	0.0390	7.655	0.1835	0.0703	159	0.79	005.05	0
005	Ash (%)	2089	7.805	0.0500	7.655	0.1835	0.0703	159	0.82	005.00	0
005	Ash (%)	0921	7.805	0.0900	7.655	0.1835	0.0703	159	0.82	005.05	0
005	Ash (%)	0353	7.810	0.0600	7.655	0.1835	0.0703	159	0.85	005.00	0
005	Ash (%)	2187	7.810	0.0400	7.655	0.1835	0.0703	159	0.85	005.03	0
005	Ash (%)	0590	7.810	0.0200	7.655	0.1835	0.0703	159	0.85	005.05	0
005	Ash (%)	2181	7.815	0.0100	7.655	0.1835	0.0703	159	0.87	005.00	0
005	Ash (%)	2124	7.820	0.1000	7.655	0.1835	0.0703	159	0.90	005.00	0
005	Ash (%)	0504	7.825	0.0300	7.655	0.1835	0.0703	159	0.93	005.00	0
005	Ash (%)	0190	7.830	0.1600	7.655	0.1835	0.0703	159	0.95	005.05	0
005	Ash (%)	0309	7.845	0.0300	7.655	0.1835	0.0703	159	1.04	005.00	0
005	Ash (%)	2193	7.845	0.0500	7.655	0.1835	0.0703	159	1.04	005.00	0
005	Ash (%)	2018	7.850	0.0200	7.655	0.1835	0.0703	159	1.06	005.05	0
005	Ash (%)	2126	7.850	0.0200	7.655	0.1835	0.0703	159	1.06	005.05	0
005	Ash (%)	0652	7.850	0.1000	7.655	0.1835	0.0703	159	1.06	005.99	0
005	Ash (%)	0613	7.860	0.2600	7.655	0.1835	0.0703	159	1.12	005.99	0
005	Ash (%)	0941	7.865	0.3100	7.655	0.1835	0.0703	159	1.15	005.05	0
005	Ash (%)	0619	7.870	0.0200	7.655	0.1835	0.0703	159	1.17	005.05	0
005	Ash (%)	0683	7.880	0.3000	7.655	0.1835	0.0703	159	1.23	005.05	0
005	Ash (%)	2192	7.880	0.0600	7.655	0.1835	0.0703	159	1.23	005.05	0
005	Ash (%)	0065	7.885	0.0500	7.655	0.1835	0.0703	159	1.25	005.05	0
005	Ash (%)	0610	7.905	0.0700	7.655	0.1835	0.0703	159	1.36	005.05	0
005	Ash (%)	2190	7.945	0.0300	7.655	0.1835	0.0703	159	1.58	005.05	0
005	Ash (%)	2129	7.955	0.1100	7.655	0.1835	0.0703	159	1.64	005.05	0
005	Ash (%)	0100	8.015	0.0300	7.655	0.1835	0.0703	159	1.96	005.00	0
005	Ash (%)	0852	8.220	0.0200	7.655	0.1835	0.0703	159	3.08	005.00	0
005	Ash (%)	0357	8.950	0.1000	7.655	0.1835	0.0703	159	7.06	005.00	0
005	Ash (%)	0226	9.275	0.0500	7.655	0.1835	0.0703	159	8.83	005.00	0
005	Ash (%)	0062	9.452	0.0310	7.655	0.1835	0.0703	159	9.79	005.00	0
005	Ash (%)	0695	7.535	0.4500	7.655	0.1835	0.0703	159	-0.65	005.05	1
005	Ash (%)	0910	7.800	0.6000	7.655	0.1835	0.0703	159	0.79	005.99	1
005	Ash (%)	0939	8.100	0.4000	7.655	0.1835	0.0703	159	2.43	005.05	1
006	Total Sugars (%)	0921	1.940	0.0000				3		006.00	0
006	Total Sugars (%)	0910	2.855	0.0100				3		006.00	0
006	Total Sugars (%)	0956	3.600	0.0000				3		006.99	0
006	Total Sugars (%)	0407	4.551	0.2090				3		006.00	1
008	Fiber, Acid Detergent (%)	0878	4.010	0.3600	5.605	1.004	0.2284	62	-1.59	008.08	0
008	Fiber, Acid Detergent (%)	0956	4.250	0.1000	5.605	1.004	0.2284	62	-1.35	008.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
008	Fiber, Acid Detergent (%)	0676	4.280	0.9800	5.605	1.004	0.2284	62	-1.32	008.02	0
008	Fiber, Acid Detergent (%)	0910	4.335	0.5300	5.605	1.004	0.2284	62	-1.26	008.08	0
008	Fiber, Acid Detergent (%)	2115	4.360	0.1400	5.605	1.004	0.2284	62	-1.24	008.08	0
008	Fiber, Acid Detergent (%)	0004	4.400	0.5600	5.605	1.004	0.2284	62	-1.20	008.08	0
008	Fiber, Acid Detergent (%)	0870	4.418	0.1048	5.605	1.004	0.2284	62	-1.18	008.08	0
008	Fiber, Acid Detergent (%)	0848	4.435	0.0100	5.605	1.004	0.2284	62	-1.17	008.08	0
008	Fiber, Acid Detergent (%)	0590	4.460	0.3200	5.605	1.004	0.2284	62	-1.14	008.02	0
008	Fiber, Acid Detergent (%)	0613	4.460	0.1800	5.605	1.004	0.2284	62	-1.14	008.99	0
008	Fiber, Acid Detergent (%)	0504	4.505	0.1500	5.605	1.004	0.2284	62	-1.10	008.02	0
008	Fiber, Acid Detergent (%)	2190	4.690	0.1200	5.605	1.004	0.2284	62	-0.91	008.08	0
008	Fiber, Acid Detergent (%)	0948	4.910	0.0400	5.605	1.004	0.2284	62	-0.69	008.02	0
008	Fiber, Acid Detergent (%)	0354	5.025	0.0100	5.605	1.004	0.2284	62	-0.58	008.08	0
008	Fiber, Acid Detergent (%)	0098	5.045	0.1900	5.605	1.004	0.2284	62	-0.56	008.02	0
008	Fiber, Acid Detergent (%)	0164	5.050	0.1000	5.605	1.004	0.2284	62	-0.55	008.08	0
008	Fiber, Acid Detergent (%)	0309	5.066	0.0044	5.605	1.004	0.2284	62	-0.54	008.02	0
008	Fiber, Acid Detergent (%)	0613	5.080	0.2800	5.605	1.004	0.2284	62	-0.52	008.08	0
008	Fiber, Acid Detergent (%)	0042	5.100	0.2600	5.605	1.004	0.2284	62	-0.50	008.08	0
008	Fiber, Acid Detergent (%)	0889	5.135	0.0700	5.605	1.004	0.2284	62	-0.47	008.08	0
008	Fiber, Acid Detergent (%)	0939	5.150	0.3600	5.605	1.004	0.2284	62	-0.45	008.08	0
008	Fiber, Acid Detergent (%)	0918	5.200	0.2982	5.605	1.004	0.2284	62	-0.40	008.08	0
008	Fiber, Acid Detergent (%)	0941	5.320	0.2200	5.605	1.004	0.2284	62	-0.28	008.99	0
008	Fiber, Acid Detergent (%)	2196	5.330	0.0000	5.605	1.004	0.2284	62	-0.27	008.08	0
008	Fiber, Acid Detergent (%)	0619	5.340	0.4600	5.605	1.004	0.2284	62	-0.26	008.08	0
008	Fiber, Acid Detergent (%)	2246	5.345	0.0500	5.605	1.004	0.2284	62	-0.26	008.02	0
008	Fiber, Acid Detergent (%)	0968	5.367	0.1170	5.605	1.004	0.2284	62	-0.24	008.08	0
008	Fiber, Acid Detergent (%)	0357	5.400	0.2000	5.605	1.004	0.2284	62	-0.20	008.08	0
008	Fiber, Acid Detergent (%)	0885	5.435	0.2322	5.605	1.004	0.2284	62	-0.17	008.02	0
008	Fiber, Acid Detergent (%)	0675	5.440	0.3400	5.605	1.004	0.2284	62	-0.16	008.08	0
008	Fiber, Acid Detergent (%)	0066	5.490	0.4200	5.605	1.004	0.2284	62	-0.11	008.08	0
008	Fiber, Acid Detergent (%)	0511	5.500	0.2000	5.605	1.004	0.2284	62	-0.10	008.99	0
008	Fiber, Acid Detergent (%)	0294	5.550	0.1000	5.605	1.004	0.2284	62	-0.05	008.08	0
008	Fiber, Acid Detergent (%)	0581	5.550	0.0400	5.605	1.004	0.2284	62	-0.05	008.08	0
008	Fiber, Acid Detergent (%)	0049	5.560	0.6400	5.605	1.004	0.2284	62	-0.04	008.08	0
008	Fiber, Acid Detergent (%)	2193	5.625	0.3100	5.605	1.004	0.2284	62	0.02	008.08	0
008	Fiber, Acid Detergent (%)	0001	5.710	0.2600	5.605	1.004	0.2284	62	0.10	008.08	0
008	Fiber, Acid Detergent (%)	0358	5.710	0.1600	5.605	1.004	0.2284	62	0.10	008.08	0
008	Fiber, Acid Detergent (%)	2129	5.723	0.1110	5.605	1.004	0.2284	62	0.12	008.08	0
008	Fiber, Acid Detergent (%)	0345	5.750	0.1000	5.605	1.004	0.2284	62	0.14	008.02	0
008	Fiber, Acid Detergent (%)	0187	5.830	0.0800	5.605	1.004	0.2284	62	0.22	008.02	0
008	Fiber, Acid Detergent (%)	0675	5.840	0.5000	5.605	1.004	0.2284	62	0.23	008.02	0
008	Fiber, Acid Detergent (%)	0610	5.900	0.1000	5.605	1.004	0.2284	62	0.29	008.99	0
008	Fiber, Acid Detergent (%)	2192	5.910	0.1800	5.605	1.004	0.2284	62	0.30	008.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
008	Fiber, Acid Detergent (%)	0190	5.910	0.0200	5.605	1.004	0.2284	62	0.30	008.08	0
008	Fiber, Acid Detergent (%)	0263	6.179	0.0320	5.605	1.004	0.2284	62	0.57	008.08	0
008	Fiber, Acid Detergent (%)	0037	6.185	0.1300	5.605	1.004	0.2284	62	0.58	008.08	0
008	Fiber, Acid Detergent (%)	0407	6.261	0.2728	5.605	1.004	0.2284	62	0.65	008.08	0
008	Fiber, Acid Detergent (%)	0596	6.315	0.1700	5.605	1.004	0.2284	62	0.71	008.08	0
008	Fiber, Acid Detergent (%)	0693	6.677	0.7400	5.605	1.004	0.2284	62	1.07	008.08	0
008	Fiber, Acid Detergent (%)	0083	6.720	0.1400	5.605	1.004	0.2284	62	1.11	008.08	0
008	Fiber, Acid Detergent (%)	0226	6.735	0.0500	5.605	1.004	0.2284	62	1.13	008.02	0
008	Fiber, Acid Detergent (%)	0689	6.750	0.3000	5.605	1.004	0.2284	62	1.14	008.02	0
008	Fiber, Acid Detergent (%)	0726	6.960	0.0000	5.605	1.004	0.2284	62	1.35	008.08	0
008	Fiber, Acid Detergent (%)	0045	7.010	0.6000	5.605	1.004	0.2284	62	1.40	008.08	0
008	Fiber, Acid Detergent (%)	0646	7.060	0.0400	5.605	1.004	0.2284	62	1.45	008.08	0
008	Fiber, Acid Detergent (%)	0278	7.250	0.1000	5.605	1.004	0.2284	62	1.64	008.08	0
008	Fiber, Acid Detergent (%)	0265	7.390	0.2600	5.605	1.004	0.2284	62	1.78	008.05	0
008	Fiber, Acid Detergent (%)	2012	7.615	0.3100	5.605	1.004	0.2284	62	2.00	008.08	0
008	Fiber, Acid Detergent (%)	0353	7.645	0.3900	5.605	1.004	0.2284	62	2.03	008.02	0
008	Fiber, Acid Detergent (%)	0510	7.900	0.6000	5.605	1.004	0.2284	62	2.29	008.08	0
008	Fiber, Acid Detergent (%)	0609	9.110	0.0200	5.605	1.004	0.2284	62	3.49	008.02	0
008	Fiber, Acid Detergent (%)	0032	5.385	1.130	5.605	1.004	0.2284	62	-0.22	008.08	1
009	Fiber, Neutral Detergent (%)	0878	14.04	0.3000	17.45	1.957	0.3950	60	-1.74	009.09	0
009	Fiber, Neutral Detergent (%)	0910	15.05	0.9000	17.45	1.957	0.3950	60	-1.23	009.07	0
009	Fiber, Neutral Detergent (%)	0353	15.11	0.3200	17.45	1.957	0.3950	60	-1.20	009.07	0
009	Fiber, Neutral Detergent (%)	0407	15.19	0.4637	17.45	1.957	0.3950	60	-1.16	009.09	0
009	Fiber, Neutral Detergent (%)	0357	15.20	0.0000	17.45	1.957	0.3950	60	-1.15	009.09	0
009	Fiber, Neutral Detergent (%)	0581	15.37	0.1400	17.45	1.957	0.3950	60	-1.06	009.09	0
009	Fiber, Neutral Detergent (%)	0164	15.40	0.4000	17.45	1.957	0.3950	60	-1.05	009.09	0
009	Fiber, Neutral Detergent (%)	0278	15.60	0.2000	17.45	1.957	0.3950	60	-0.95	009.09	0
009	Fiber, Neutral Detergent (%)	0598	15.65	0.2100	17.45	1.957	0.3950	60	-0.92	009.09	0
009	Fiber, Neutral Detergent (%)	0613	15.67	0.3500	17.45	1.957	0.3950	60	-0.91	009.99	0
009	Fiber, Neutral Detergent (%)	0619	15.68	0.7600	17.45	1.957	0.3950	60	-0.91	009.09	0
009	Fiber, Neutral Detergent (%)	0870	15.70	0.6091	17.45	1.957	0.3950	60	-0.89	009.09	0
009	Fiber, Neutral Detergent (%)	0354	15.89	0.0000	17.45	1.957	0.3950	60	-0.80	009.09	0
009	Fiber, Neutral Detergent (%)	0049	15.90	0.1100	17.45	1.957	0.3950	60	-0.80	009.09	0
009	Fiber, Neutral Detergent (%)	0263	15.90	0.1762	17.45	1.957	0.3950	60	-0.79	009.09	0
009	Fiber, Neutral Detergent (%)	2190	16.01	0.0000	17.45	1.957	0.3950	60	-0.74	009.09	0
009	Fiber, Neutral Detergent (%)	0956	16.05	0.7000	17.45	1.957	0.3950	60	-0.72	009.09	0
009	Fiber, Neutral Detergent (%)	0889	16.12	0.0900	17.45	1.957	0.3950	60	-0.68	009.09	0
009	Fiber, Neutral Detergent (%)	0939	16.20	0.1200	17.45	1.957	0.3950	60	-0.64	009.09	0
009	Fiber, Neutral Detergent (%)	2193	16.23	0.0200	17.45	1.957	0.3950	60	-0.62	009.09	0
009	Fiber, Neutral Detergent (%)	0675	16.23	0.1000	17.45	1.957	0.3950	60	-0.62	009.09	0
009	Fiber, Neutral Detergent (%)	0889	16.38	0.0100	17.45	1.957	0.3950	60	-0.55	009.99	0
009	Fiber, Neutral Detergent (%)	0613	16.40	0.9800	17.45	1.957	0.3950	60	-0.54	009.09	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
009	Fiber, Neutral Detergent (%)	0504	16.54	0.0600	17.45	1.957	0.3950	60	-0.47	009.04	0
009	Fiber, Neutral Detergent (%)	0675	16.56	0.2400	17.45	1.957	0.3950	60	-0.46	009.07	0
009	Fiber, Neutral Detergent (%)	0358	16.57	0.1900	17.45	1.957	0.3950	60	-0.45	009.09	0
009	Fiber, Neutral Detergent (%)	0098	16.75	0.9000	17.45	1.957	0.3950	60	-0.36	009.07	0
009	Fiber, Neutral Detergent (%)	2196	16.86	0.0000	17.45	1.957	0.3950	60	-0.30	009.09	0
009	Fiber, Neutral Detergent (%)	0510	16.90	0.4000	17.45	1.957	0.3950	60	-0.28	009.09	0
009	Fiber, Neutral Detergent (%)	0646	17.12	1.360	17.45	1.957	0.3950	60	-0.17	009.09	0
009	Fiber, Neutral Detergent (%)	0083	17.25	0.0700	17.45	1.957	0.3950	60	-0.11	009.09	0
009	Fiber, Neutral Detergent (%)	0187	17.46	0.0400	17.45	1.957	0.3950	60	0.00	009.07	0
009	Fiber, Neutral Detergent (%)	0848	17.58	0.0000	17.45	1.957	0.3950	60	0.07	009.09	0
009	Fiber, Neutral Detergent (%)	0309	17.71	0.0257	17.45	1.957	0.3950	60	0.13	009.07	0
009	Fiber, Neutral Detergent (%)	0001	17.79	0.7800	17.45	1.957	0.3950	60	0.17	009.09	0
009	Fiber, Neutral Detergent (%)	0037	17.94	0.1300	17.45	1.957	0.3950	60	0.25	009.09	0
009	Fiber, Neutral Detergent (%)	0345	18.00	0.4000	17.45	1.957	0.3950	60	0.28	009.07	0
009	Fiber, Neutral Detergent (%)	0918	18.09	1.378	17.45	1.957	0.3950	60	0.32	009.09	0
009	Fiber, Neutral Detergent (%)	0968	18.26	0.4890	17.45	1.957	0.3950	60	0.41	009.09	0
009	Fiber, Neutral Detergent (%)	2192	18.39	0.5100	17.45	1.957	0.3950	60	0.48	009.07	0
009	Fiber, Neutral Detergent (%)	0590	18.47	1.000	17.45	1.957	0.3950	60	0.52	009.07	0
009	Fiber, Neutral Detergent (%)	2129	18.52	1.100	17.45	1.957	0.3950	60	0.54	009.09	0
009	Fiber, Neutral Detergent (%)	0032	18.53	0.0400	17.45	1.957	0.3950	60	0.55	009.09	0
009	Fiber, Neutral Detergent (%)	0693	18.68	0.5530	17.45	1.957	0.3950	60	0.63	009.09	0
009	Fiber, Neutral Detergent (%)	0596	18.72	0.5500	17.45	1.957	0.3950	60	0.65	009.09	0
009	Fiber, Neutral Detergent (%)	0948	18.81	0.4500	17.45	1.957	0.3950	60	0.69	009.09	0
009	Fiber, Neutral Detergent (%)	0676	19.01	1.020	17.45	1.957	0.3950	60	0.80	009.07	0
009	Fiber, Neutral Detergent (%)	0066	19.12	0.3400	17.45	1.957	0.3950	60	0.85	009.09	0
009	Fiber, Neutral Detergent (%)	0885	19.19	0.2548	17.45	1.957	0.3950	60	0.89	009.07	0
009	Fiber, Neutral Detergent (%)	0638	19.20	0.4000	17.45	1.957	0.3950	60	0.89	009.09	0
009	Fiber, Neutral Detergent (%)	0294	19.30	0.2000	17.45	1.957	0.3950	60	0.94	009.09	0
009	Fiber, Neutral Detergent (%)	0226	19.56	0.0800	17.45	1.957	0.3950	60	1.08	009.07	0
009	Fiber, Neutral Detergent (%)	0689	20.10	0.4000	17.45	1.957	0.3950	60	1.35	009.07	0
009	Fiber, Neutral Detergent (%)	2012	20.44	0.3000	17.45	1.957	0.3950	60	1.53	009.09	0
009	Fiber, Neutral Detergent (%)	0265	20.60	0.2000	17.45	1.957	0.3950	60	1.61	009.09	0
009	Fiber, Neutral Detergent (%)	2115	20.66	0.4300	17.45	1.957	0.3950	60	1.64	009.09	0
009	Fiber, Neutral Detergent (%)	0045	21.00	0.8000	17.45	1.957	0.3950	60	1.81	009.09	0
009	Fiber, Neutral Detergent (%)	0726	21.45	0.0800	17.45	1.957	0.3950	60	2.04	009.09	0
009	Fiber, Neutral Detergent (%)	0190	21.83	1.140	17.45	1.957	0.3950	60	2.24	009.09	0
009	Fiber, Neutral Detergent (%)	0609	22.63	0.4300	17.45	1.957	0.3950	60	2.64	009.07	0
009	Fiber, Neutral Detergent (%)	2246	11.76	18.20	17.45	1.957	0.3950	60	-2.91	009.07	1
009	Fiber, Neutral Detergent (%)	2246	34.47	0.0700	17.45	1.957	0.3950	60	8.69	009.04	2
010	Moisture (%)	0546	6.275	0.1100	7.468	0.4730	0.1095	22	-2.52	010.99	0
010	Moisture (%)	2128	6.465	0.1765	7.468	0.4730	0.1095	22	-2.12	010.11	0
010	Moisture (%)	0726	6.955	0.0300	7.468	0.4730	0.1095	22	-1.08	010.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
010	Moisture (%)	2178	7.175	0.2100	7.468	0.4730	0.1095	22	-0.62	010.99	0
010	Moisture (%)	0948	7.185	0.0500	7.468	0.4730	0.1095	22	-0.60	010.99	0
010	Moisture (%)	0964	7.185	0.0620	7.468	0.4730	0.1095	22	-0.60	010.99	0
010	Moisture (%)	0843	7.210	0.2200	7.468	0.4730	0.1095	22	-0.55	010.03	0
010	Moisture (%)	0884	7.300	0.2000	7.468	0.4730	0.1095	22	-0.36	010.99	0
010	Moisture (%)	0164	7.330	0.0000	7.468	0.4730	0.1095	22	-0.29	010.03	0
010	Moisture (%)	2043	7.395	0.2300	7.468	0.4730	0.1095	22	-0.15	010.99	0
010	Moisture (%)	0028	7.400	0.0200	7.468	0.4730	0.1095	22	-0.14	010.99	0
010	Moisture (%)	0939	7.400	0.0800	7.468	0.4730	0.1095	22	-0.14	010.99	0
010	Moisture (%)	0852	7.445	0.1300	7.468	0.4730	0.1095	22	-0.05	010.11	0
010	Moisture (%)	0652	7.450	0.3000	7.468	0.4730	0.1095	22	-0.04	010.99	0
010	Moisture (%)	0921	7.545	0.0100	7.468	0.4730	0.1095	22	0.16	010.99	0
010	Moisture (%)	2076	7.780	0.1200	7.468	0.4730	0.1095	22	0.66	010.99	0
010	Moisture (%)	0952	7.800	0.0200	7.468	0.4730	0.1095	22	0.70	010.99	0
010	Moisture (%)	0190	7.830	0.0800	7.468	0.4730	0.1095	22	0.77	010.99	0
010	Moisture (%)	0914	7.860	0.0600	7.468	0.4730	0.1095	22	0.83	010.99	0
010	Moisture (%)	0889	8.295	0.0300	7.468	0.4730	0.1095	22	1.75	010.11	0
010	Moisture (%)	0405	9.120	0.0000	7.468	0.4730	0.1095	22	3.49	010.99	0
010	Moisture (%)	2150	9.655	0.2700	7.468	0.4730	0.1095	22	4.62	010.99	0
010	Moisture (%)	2089	7.275	0.5500	7.468	0.4730	0.1095	22	-0.41	010.99	1
010	Moisture (%)	2129	92.98	0.1100	7.468	0.4730	0.1095	22	180.78	010.99	2
011	Loss on Drying (%)	2016	7.050	0.0800	7.834	0.2573	0.0927	76	-3.05	011.01	0
011	Loss on Drying (%)	2031	7.322	0.0047	7.834	0.2573	0.0927	76	-1.99	011.01	0
011	Loss on Drying (%)	0874	7.325	0.0100	7.834	0.2573	0.0927	76	-1.98	011.01	0
011	Loss on Drying (%)	0650	7.355	0.2900	7.834	0.2573	0.0927	76	-1.86	011.01	0
011	Loss on Drying (%)	2198	7.395	0.0300	7.834	0.2573	0.0927	76	-1.71	011.01	0
011	Loss on Drying (%)	0554	7.400	0.0800	7.834	0.2573	0.0927	76	-1.69	011.01	0
011	Loss on Drying (%)	0643	7.400	0.4000	7.834	0.2573	0.0927	76	-1.69	011.01	0
011	Loss on Drying (%)	0298	7.420	0.0200	7.834	0.2573	0.0927	76	-1.61	011.01	0
011	Loss on Drying (%)	0675	7.430	0.0200	7.834	0.2573	0.0927	76	-1.57	011.01	0
011	Loss on Drying (%)	0881	7.440	0.0470	7.834	0.2573	0.0927	76	-1.53	011.01	0
011	Loss on Drying (%)	0047	7.450	0.3000	7.834	0.2573	0.0927	76	-1.49	011.01	0
011	Loss on Drying (%)	2162	7.475	0.0500	7.834	0.2573	0.0927	76	-1.39	011.01	0
011	Loss on Drying (%)	2181	7.490	0.2000	7.834	0.2573	0.0927	76	-1.34	011.01	0
011	Loss on Drying (%)	0062	7.523	0.0060	7.834	0.2573	0.0927	76	-1.21	011.01	0
011	Loss on Drying (%)	0598	7.530	0.0000	7.834	0.2573	0.0927	76	-1.18	011.01	0
011	Loss on Drying (%)	2083	7.590	0.3800	7.834	0.2573	0.0927	76	-0.95	011.02	0
011	Loss on Drying (%)	0857	7.600	0.0000	7.834	0.2573	0.0927	76	-0.91	011.99	0
011	Loss on Drying (%)	2168	7.670	0.1400	7.834	0.2573	0.0927	76	-0.64	011.01	0
011	Loss on Drying (%)	2006	7.745	0.0300	7.834	0.2573	0.0927	76	-0.34	011.01	0
011	Loss on Drying (%)	2033	7.745	0.0100	7.834	0.2573	0.0927	76	-0.34	011.01	0
011	Loss on Drying (%)	0171	7.760	0.0200	7.834	0.2573	0.0927	76	-0.29	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
011	Loss on Drying (%)	0208	7.770	0.0000	7.834	0.2573	0.0927	76	-0.25	011.01	0
011	Loss on Drying (%)	0100	7.775	0.0500	7.834	0.2573	0.0927	76	-0.23	011.01	0
011	Loss on Drying (%)	0226	7.775	0.1100	7.834	0.2573	0.0927	76	-0.23	011.01	0
011	Loss on Drying (%)	0294	7.800	0.0000	7.834	0.2573	0.0927	76	-0.13	011.01	0
011	Loss on Drying (%)	2196	7.800	0.0000	7.834	0.2573	0.0927	76	-0.13	011.01	0
011	Loss on Drying (%)	0541	7.805	0.0100	7.834	0.2573	0.0927	76	-0.11	011.01	0
011	Loss on Drying (%)	0674	7.810	0.0000	7.834	0.2573	0.0927	76	-0.09	011.01	0
011	Loss on Drying (%)	0953	7.830	0.1600	7.834	0.2573	0.0927	76	-0.01	011.01	0
011	Loss on Drying (%)	0242	7.835	0.0100	7.834	0.2573	0.0927	76	0.01	011.01	0
011	Loss on Drying (%)	0653	7.835	0.0300	7.834	0.2573	0.0927	76	0.01	011.01	0
011	Loss on Drying (%)	2073	7.835	0.0100	7.834	0.2573	0.0927	76	0.01	011.01	0
011	Loss on Drying (%)	0682	7.840	0.0000	7.834	0.2573	0.0927	76	0.02	011.01	0
011	Loss on Drying (%)	0511	7.845	0.3300	7.834	0.2573	0.0927	76	0.04	011.01	0
011	Loss on Drying (%)	0144	7.850	0.1600	7.834	0.2573	0.0927	76	0.06	011.01	0
011	Loss on Drying (%)	0520	7.850	0.1000	7.834	0.2573	0.0927	76	0.06	011.01	0
011	Loss on Drying (%)	0968	7.854	0.0060	7.834	0.2573	0.0927	76	0.08	011.01	0
011	Loss on Drying (%)	2193	7.855	0.2500	7.834	0.2573	0.0927	76	0.08	011.01	0
011	Loss on Drying (%)	0417	7.865	0.2300	7.834	0.2573	0.0927	76	0.12	011.02	0
011	Loss on Drying (%)	0843	7.890	0.1200	7.834	0.2573	0.0927	76	0.22	011.02	0
011	Loss on Drying (%)	0051	7.895	0.2300	7.834	0.2573	0.0927	76	0.24	011.01	0
011	Loss on Drying (%)	2184	7.895	0.1300	7.834	0.2573	0.0927	76	0.24	011.01	0
011	Loss on Drying (%)	0942	7.895	0.2900	7.834	0.2573	0.0927	76	0.24	011.02	0
011	Loss on Drying (%)	0646	7.900	0.0400	7.834	0.2573	0.0927	76	0.26	011.01	0
011	Loss on Drying (%)	0596	7.905	0.1500	7.834	0.2573	0.0927	76	0.28	011.01	0
011	Loss on Drying (%)	0859	7.914	0.0030	7.834	0.2573	0.0927	76	0.31	011.01	0
011	Loss on Drying (%)	0425	7.915	0.0900	7.834	0.2573	0.0927	76	0.32	011.01	0
011	Loss on Drying (%)	0860	7.920	0.0000	7.834	0.2573	0.0927	76	0.34	011.01	0
011	Loss on Drying (%)	0958	7.920	0.0400	7.834	0.2573	0.0927	76	0.34	011.01	0
011	Loss on Drying (%)	0309	7.935	0.0100	7.834	0.2573	0.0927	76	0.39	011.01	0
011	Loss on Drying (%)	2022	7.935	0.0300	7.834	0.2573	0.0927	76	0.39	011.01	0
011	Loss on Drying (%)	0870	7.936	0.1120	7.834	0.2573	0.0927	76	0.40	011.01	0
011	Loss on Drying (%)	0265	7.950	0.3000	7.834	0.2573	0.0927	76	0.45	011.01	0
011	Loss on Drying (%)	0108	7.960	0.0200	7.834	0.2573	0.0927	76	0.49	011.01	0
011	Loss on Drying (%)	0229	7.961	0.0227	7.834	0.2573	0.0927	76	0.50	011.01	0
011	Loss on Drying (%)	0098	7.965	0.0700	7.834	0.2573	0.0927	76	0.51	011.01	0
011	Loss on Drying (%)	0539	7.970	0.1000	7.834	0.2573	0.0927	76	0.53	011.01	0
011	Loss on Drying (%)	0263	7.995	0.0500	7.834	0.2573	0.0927	76	0.63	011.01	0
011	Loss on Drying (%)	0651	8.004	0.0780	7.834	0.2573	0.0927	76	0.66	011.01	0
011	Loss on Drying (%)	2246	8.005	0.0100	7.834	0.2573	0.0927	76	0.67	011.01	0
011	Loss on Drying (%)	0164	8.010	0.0600	7.834	0.2573	0.0927	76	0.69	011.01	0
011	Loss on Drying (%)	0957	8.015	0.1100	7.834	0.2573	0.0927	76	0.70	011.01	0
011	Loss on Drying (%)	0202	8.020	0.0130	7.834	0.2573	0.0927	76	0.72	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
011	Loss on Drying (%)	0955	8.025	0.1300	7.834	0.2573	0.0927	76	0.74	011.01	0
011	Loss on Drying (%)	0354	8.030	0.0000	7.834	0.2573	0.0927	76	0.76	011.01	0
011	Loss on Drying (%)	2207	8.050	0.1000	7.834	0.2573	0.0927	76	0.84	011.01	0
011	Loss on Drying (%)	0066	8.060	0.1000	7.834	0.2573	0.0927	76	0.88	011.01	0
011	Loss on Drying (%)	0205	8.085	0.1070	7.834	0.2573	0.0927	76	0.97	011.01	0
011	Loss on Drying (%)	0589	8.135	0.0500	7.834	0.2573	0.0927	76	1.17	011.01	0
011	Loss on Drying (%)	2188	8.135	0.0100	7.834	0.2573	0.0927	76	1.17	011.01	0
011	Loss on Drying (%)	2242	8.150	0.1000	7.834	0.2573	0.0927	76	1.23	011.01	0
011	Loss on Drying (%)	0004	8.260	0.3800	7.834	0.2573	0.0927	76	1.66	011.01	0
011	Loss on Drying (%)	0407	8.267	0.0451	7.834	0.2573	0.0927	76	1.69	011.01	0
011	Loss on Drying (%)	2058	8.290	0.1000	7.834	0.2573	0.0927	76	1.77	011.01	0
011	Loss on Drying (%)	0358	8.325	0.1900	7.834	0.2573	0.0927	76	1.91	011.01	0
011	Loss on Drying (%)	0852	8.520	0.0800	7.834	0.2573	0.0927	76	2.67	011.99	0
011	Loss on Drying (%)	0510	92.25	0.1000	7.834	0.2573	0.0927	76	328.09	011.01	2
012	Starch (%)	0918	23.97	1.460	32.70	1.131	0.3744	40	-7.72	012.01	0
012	Starch (%)	0609	28.97	0.0000	32.70	1.131	0.3744	40	-3.29	012.00	0
012	Starch (%)	0889	29.46	0.1300	32.70	1.131	0.3744	40	-2.86	012.11	0
012	Starch (%)	0676	30.17	0.2800	32.70	1.131	0.3744	40	-2.23	012.01	0
012	Starch (%)	0353	31.17	0.1200	32.70	1.131	0.3744	40	-1.35	012.04	0
012	Starch (%)	0870	31.44	0.6841	32.70	1.131	0.3744	40	-1.11	012.01	0
012	Starch (%)	0164	31.60	0.2000	32.70	1.131	0.3744	40	-0.97	012.01	0
012	Starch (%)	0510	31.60	0.6000	32.70	1.131	0.3744	40	-0.97	012.04	0
012	Starch (%)	2242	31.71	0.1200	32.70	1.131	0.3744	40	-0.87	012.01	0
012	Starch (%)	0885	31.80	0.4532	32.70	1.131	0.3744	40	-0.80	012.01	0
012	Starch (%)	2188	31.97	0.3500	32.70	1.131	0.3744	40	-0.65	012.00	0
012	Starch (%)	0638	32.20	0.0000	32.70	1.131	0.3744	40	-0.44	012.00	0
012	Starch (%)	0045	32.20	1.400	32.70	1.131	0.3744	40	-0.44	012.01	0
012	Starch (%)	2193	32.37	0.3600	32.70	1.131	0.3744	40	-0.29	012.00	0
012	Starch (%)	2196	32.40	0.0000	32.70	1.131	0.3744	40	-0.26	012.00	0
012	Starch (%)	2181	32.50	0.2200	32.70	1.131	0.3744	40	-0.17	012.00	0
012	Starch (%)	0265	32.60	0.4000	32.70	1.131	0.3744	40	-0.08	012.01	0
012	Starch (%)	0956	32.60	1.600	32.70	1.131	0.3744	40	-0.08	012.01	0
012	Starch (%)	0407	32.62	0.2900	32.70	1.131	0.3744	40	-0.07	012.03	0
012	Starch (%)	0689	32.75	0.5000	32.70	1.131	0.3744	40	0.05	012.00	0
012	Starch (%)	0098	32.75	0.1000	32.70	1.131	0.3744	40	0.05	012.03	0
012	Starch (%)	0910	32.95	0.3000	32.70	1.131	0.3744	40	0.23	012.00	0
012	Starch (%)	0948	32.99	0.2100	32.70	1.131	0.3744	40	0.26	012.04	0
012	Starch (%)	0955	33.00	1.800	32.70	1.131	0.3744	40	0.27	012.03	0
012	Starch (%)	2006	33.09	0.0200	32.70	1.131	0.3744	40	0.35	012.00	0
012	Starch (%)	0278	33.15	0.1000	32.70	1.131	0.3744	40	0.40	012.04	0
012	Starch (%)	2192	33.18	0.1700	32.70	1.131	0.3744	40	0.42	012.00	0
012	Starch (%)	2009	33.32	0.0310	32.70	1.131	0.3744	40	0.55	012.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
012	Starch (%)	0693	33.36	0.7530	32.70	1.131	0.3744	40	0.58	012.01	0
012	Starch (%)	0066	33.50	0.4100	32.70	1.131	0.3744	40	0.71	012.03	0
012	Starch (%)	0164	33.55	0.5000	32.70	1.131	0.3744	40	0.76	012.00	0
012	Starch (%)	0610	33.60	0.2000	32.70	1.131	0.3744	40	0.80	012.00	0
012	Starch (%)	0941	33.66	0.5100	32.70	1.131	0.3744	40	0.85	012.00	0
012	Starch (%)	0939	33.70	0.0000	32.70	1.131	0.3744	40	0.89	012.00	0
012	Starch (%)	0354	33.73	0.1000	32.70	1.131	0.3744	40	0.91	012.00	0
012	Starch (%)	0619	33.80	0.2000	32.70	1.131	0.3744	40	0.98	012.00	0
012	Starch (%)	0683	33.86	0.2200	32.70	1.131	0.3744	40	1.03	012.00	0
012	Starch (%)	0878	34.35	0.1000	32.70	1.131	0.3744	40	1.46	012.00	0
012	Starch (%)	2128	34.93	0.0860	32.70	1.131	0.3744	40	1.97	012.11	0
012	Starch (%)	2089	37.23	0.0000	32.70	1.131	0.3744	40	4.01	012.01	0
012	Starch (%)	0613	37.26	13.64	32.70	1.131	0.3744	40	4.03	012.01	1
012	Starch (%)	2124	14.45	0.1000	32.70	1.131	0.3744	40	-16.13	012.03	2
013	Fat, Acid Pretreat (%)	0554	6.230	0.4400	9.045	0.6459	0.1763	84	-4.36	013.10	0
013	Fat, Acid Pretreat (%)	0353	7.360	0.3600	9.045	0.6459	0.1763	84	-2.61	013.10	0
013	Fat, Acid Pretreat (%)	2150	7.470	0.2000	9.045	0.6459	0.1763	84	-2.44	013.00	0
013	Fat, Acid Pretreat (%)	0263	7.685	0.0440	9.045	0.6459	0.1763	84	-2.11	013.10	0
013	Fat, Acid Pretreat (%)	0613	7.860	0.0800	9.045	0.6459	0.1763	84	-1.83	013.12	0
013	Fat, Acid Pretreat (%)	2162	7.915	0.2300	9.045	0.6459	0.1763	84	-1.75	013.10	0
013	Fat, Acid Pretreat (%)	0876	8.148	0.6370	9.045	0.6459	0.1763	84	-1.39	013.13	0
013	Fat, Acid Pretreat (%)	2124	8.210	0.0000	9.045	0.6459	0.1763	84	-1.29	013.10	0
013	Fat, Acid Pretreat (%)	0505	8.265	0.1100	9.045	0.6459	0.1763	84	-1.21	013.13	0
013	Fat, Acid Pretreat (%)	0689	8.300	0.2000	9.045	0.6459	0.1763	84	-1.15	013.00	0
013	Fat, Acid Pretreat (%)	0873	8.305	0.1300	9.045	0.6459	0.1763	84	-1.15	013.02	0
013	Fat, Acid Pretreat (%)	0638	8.350	0.1000	9.045	0.6459	0.1763	84	-1.08	013.10	0
013	Fat, Acid Pretreat (%)	0011	8.370	0.2200	9.045	0.6459	0.1763	84	-1.04	013.02	0
013	Fat, Acid Pretreat (%)	0845	8.370	0.1400	9.045	0.6459	0.1763	84	-1.04	013.10	0
013	Fat, Acid Pretreat (%)	0853	8.400	0.6000	9.045	0.6459	0.1763	84	-1.00	013.02	0
013	Fat, Acid Pretreat (%)	0083	8.430	0.0800	9.045	0.6459	0.1763	84	-0.95	013.00	0
013	Fat, Acid Pretreat (%)	0009	8.465	0.3900	9.045	0.6459	0.1763	84	-0.90	013.13	0
013	Fat, Acid Pretreat (%)	2006	8.525	0.1100	9.045	0.6459	0.1763	84	-0.80	013.10	0
013	Fat, Acid Pretreat (%)	0096	8.538	0.3299	9.045	0.6459	0.1763	84	-0.79	013.00	0
013	Fat, Acid Pretreat (%)	0581	8.560	0.1000	9.045	0.6459	0.1763	84	-0.75	013.13	0
013	Fat, Acid Pretreat (%)	0019	8.590	0.0400	9.045	0.6459	0.1763	84	-0.70	013.02	0
013	Fat, Acid Pretreat (%)	2146	8.610	0.1800	9.045	0.6459	0.1763	84	-0.67	013.00	0
013	Fat, Acid Pretreat (%)	2156	8.625	0.0900	9.045	0.6459	0.1763	84	-0.65	013.00	0
013	Fat, Acid Pretreat (%)	0036	8.642	0.0532	9.045	0.6459	0.1763	84	-0.62	013.10	0
013	Fat, Acid Pretreat (%)	0650	8.685	0.1500	9.045	0.6459	0.1763	84	-0.56	013.02	0
013	Fat, Acid Pretreat (%)	2187	8.690	0.2000	9.045	0.6459	0.1763	84	-0.55	013.00	0
013	Fat, Acid Pretreat (%)	0027	8.693	0.1280	9.045	0.6459	0.1763	84	-0.54	013.13	0
013	Fat, Acid Pretreat (%)	0407	8.735	0.0900	9.045	0.6459	0.1763	84	-0.48	013.13	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
013	Fat, Acid Pretreat (%)	2073	8.745	0.0500	9.045	0.6459	0.1763	84	-0.46	013.00	0
013	Fat, Acid Pretreat (%)	0510	8.750	0.3000	9.045	0.6459	0.1763	84	-0.46	013.02	0
013	Fat, Acid Pretreat (%)	0265	8.770	0.1000	9.045	0.6459	0.1763	84	-0.43	013.13	0
013	Fat, Acid Pretreat (%)	0610	8.820	0.0000	9.045	0.6459	0.1763	84	-0.35	013.10	0
013	Fat, Acid Pretreat (%)	2058	8.860	0.0800	9.045	0.6459	0.1763	84	-0.29	013.00	0
013	Fat, Acid Pretreat (%)	2196	8.880	0.0000	9.045	0.6459	0.1763	84	-0.25	013.10	0
013	Fat, Acid Pretreat (%)	0033	8.895	0.1100	9.045	0.6459	0.1763	84	-0.23	013.10	0
013	Fat, Acid Pretreat (%)	0682	8.900	0.0000	9.045	0.6459	0.1763	84	-0.22	013.02	0
013	Fat, Acid Pretreat (%)	0074	8.905	0.1300	9.045	0.6459	0.1763	84	-0.22	013.02	0
013	Fat, Acid Pretreat (%)	0843	8.905	0.4700	9.045	0.6459	0.1763	84	-0.22	013.13	0
013	Fat, Acid Pretreat (%)	0541	8.945	0.1300	9.045	0.6459	0.1763	84	-0.15	013.00	0
013	Fat, Acid Pretreat (%)	0910	8.950	0.1800	9.045	0.6459	0.1763	84	-0.15	013.00	0
013	Fat, Acid Pretreat (%)	0242	8.995	0.9100	9.045	0.6459	0.1763	84	-0.08	013.00	0
013	Fat, Acid Pretreat (%)	2058	9.020	0.1400	9.045	0.6459	0.1763	84	-0.04	013.02	0
013	Fat, Acid Pretreat (%)	0653	9.090	0.1400	9.045	0.6459	0.1763	84	0.07	013.13	0
013	Fat, Acid Pretreat (%)	0884	9.150	0.3000	9.045	0.6459	0.1763	84	0.16	013.00	0
013	Fat, Acid Pretreat (%)	0035	9.210	0.0000	9.045	0.6459	0.1763	84	0.26	013.00	0
013	Fat, Acid Pretreat (%)	2259	9.215	0.1100	9.045	0.6459	0.1763	84	0.26	013.02	0
013	Fat, Acid Pretreat (%)	0208	9.220	0.2400	9.045	0.6459	0.1763	84	0.27	013.02	0
013	Fat, Acid Pretreat (%)	0035	9.220	0.1000	9.045	0.6459	0.1763	84	0.27	013.13	0
013	Fat, Acid Pretreat (%)	0229	9.230	0.0800	9.045	0.6459	0.1763	84	0.29	013.00	0
013	Fat, Acid Pretreat (%)	0652	9.250	0.1000	9.045	0.6459	0.1763	84	0.32	013.00	0
013	Fat, Acid Pretreat (%)	0407	9.268	0.1567	9.045	0.6459	0.1763	84	0.35	013.00	0
013	Fat, Acid Pretreat (%)	0309	9.274	0.1025	9.045	0.6459	0.1763	84	0.35	013.00	0
013	Fat, Acid Pretreat (%)	2252	9.280	0.5200	9.045	0.6459	0.1763	84	0.36	013.00	0
013	Fat, Acid Pretreat (%)	0100	9.300	0.3800	9.045	0.6459	0.1763	84	0.40	013.02	0
013	Fat, Acid Pretreat (%)	0948	9.305	0.0500	9.045	0.6459	0.1763	84	0.40	013.00	0
013	Fat, Acid Pretreat (%)	0354	9.375	0.0300	9.045	0.6459	0.1763	84	0.51	013.02	0
013	Fat, Acid Pretreat (%)	0098	9.400	0.1000	9.045	0.6459	0.1763	84	0.55	013.02	0
013	Fat, Acid Pretreat (%)	2053	9.410	0.0400	9.045	0.6459	0.1763	84	0.57	013.00	0
013	Fat, Acid Pretreat (%)	2033	9.435	0.2500	9.045	0.6459	0.1763	84	0.60	013.02	0
013	Fat, Acid Pretreat (%)	0958	9.480	0.1400	9.045	0.6459	0.1763	84	0.67	013.02	0
013	Fat, Acid Pretreat (%)	0065	9.485	0.1900	9.045	0.6459	0.1763	84	0.68	013.02	0
013	Fat, Acid Pretreat (%)	2219	9.495	0.2700	9.045	0.6459	0.1763	84	0.70	013.00	0
013	Fat, Acid Pretreat (%)	2220	9.515	0.2300	9.045	0.6459	0.1763	84	0.73	013.00	0
013	Fat, Acid Pretreat (%)	0957	9.515	0.0900	9.045	0.6459	0.1763	84	0.73	013.02	0
013	Fat, Acid Pretreat (%)	0139	9.540	0.0800	9.045	0.6459	0.1763	84	0.77	013.00	0
013	Fat, Acid Pretreat (%)	0870	9.548	0.0815	9.045	0.6459	0.1763	84	0.78	013.02	0
013	Fat, Acid Pretreat (%)	2016	9.560	0.2200	9.045	0.6459	0.1763	84	0.80	013.00	0
013	Fat, Acid Pretreat (%)	0955	9.575	0.0100	9.045	0.6459	0.1763	84	0.82	013.13	0
013	Fat, Acid Pretreat (%)	2076	9.589	0.2737	9.045	0.6459	0.1763	84	0.84	013.00	0
013	Fat, Acid Pretreat (%)	0202	9.605	0.0230	9.045	0.6459	0.1763	84	0.87	013.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
013	Fat, Acid Pretreat (%)	0675	9.610	0.1200	9.045	0.6459	0.1763	84	0.88	013.02	0
013	Fat, Acid Pretreat (%)	0013	9.625	0.1300	9.045	0.6459	0.1763	84	0.90	013.02	0
013	Fat, Acid Pretreat (%)	0676	9.670	0.0200	9.045	0.6459	0.1763	84	0.97	013.02	0
013	Fat, Acid Pretreat (%)	2089	9.705	0.0500	9.045	0.6459	0.1763	84	1.02	013.00	0
013	Fat, Acid Pretreat (%)	2128	9.720	0.0200	9.045	0.6459	0.1763	84	1.05	013.12	0
013	Fat, Acid Pretreat (%)	0019	9.810	0.1000	9.045	0.6459	0.1763	84	1.18	013.13	0
013	Fat, Acid Pretreat (%)	0003	9.855	0.4900	9.045	0.6459	0.1763	84	1.25	013.13	0
013	Fat, Acid Pretreat (%)	0004	9.900	0.0800	9.045	0.6459	0.1763	84	1.32	013.02	0
013	Fat, Acid Pretreat (%)	0017	9.950	0.1000	9.045	0.6459	0.1763	84	1.40	013.02	0
013	Fat, Acid Pretreat (%)	0643	10.01	0.0600	9.045	0.6459	0.1763	84	1.49	013.02	0
013	Fat, Acid Pretreat (%)	0171	10.11	0.1400	9.045	0.6459	0.1763	84	1.65	013.02	0
013	Fat, Acid Pretreat (%)	0013	10.39	0.0900	9.045	0.6459	0.1763	84	2.08	013.13	0
013	Fat, Acid Pretreat (%)	0425	10.53	0.6600	9.045	0.6459	0.1763	84	2.30	013.02	0
013	Fat, Acid Pretreat (%)	0874	10.96	0.4100	9.045	0.6459	0.1763	84	2.96	013.02	0
013	Fat, Acid Pretreat (%)	0051	7.350	1.100	9.045	0.6459	0.1763	84	-2.62	013.02	1
013	Fat, Acid Pretreat (%)	0511	8.405	1.150	9.045	0.6459	0.1763	84	-0.99	013.02	1
015	Aluminum (mg / kg (ppm))	0186	32.10	0.8000	62.42	11.18	2.036	13	-2.71	015.52	0
015	Aluminum (mg / kg (ppm))	0049	47.99	1.600	62.42	11.18	2.036	13	-1.29	015.41	0
015	Aluminum (mg / kg (ppm))	0918	55.65	2.792	62.42	11.18	2.036	13	-0.61	015.43	0
015	Aluminum (mg / kg (ppm))	0510	57.00	0.0000	62.42	11.18	2.036	13	-0.49	015.43	0
015	Aluminum (mg / kg (ppm))	0560	57.15	4.300	62.42	11.18	2.036	13	-0.47	015.42	0
015	Aluminum (mg / kg (ppm))	0171	61.48	0.4200	62.42	11.18	2.036	13	-0.08	015.41	0
015	Aluminum (mg / kg (ppm))	0169	62.60	1.800	62.42	11.18	2.036	13	0.02	015.43	0
015	Aluminum (mg / kg (ppm))	0407	64.83	0.5041	62.42	11.18	2.036	13	0.21	015.41	0
015	Aluminum (mg / kg (ppm))	0164	66.50	5.000	62.42	11.18	2.036	13	0.36	015.41	0
015	Aluminum (mg / kg (ppm))	0425	67.13	0.7500	62.42	11.18	2.036	13	0.42	015.43	0
015	Aluminum (mg / kg (ppm))	0345	70.35	3.500	62.42	11.18	2.036	13	0.71	015.43	0
015	Aluminum (mg / kg (ppm))	0520	76.00	0.0000	62.42	11.18	2.036	13	1.21	015.41	0
015	Aluminum (mg / kg (ppm))	0353	86.50	5.000	62.42	11.18	2.036	13	2.15	015.43	0
017	Boron (mg / kg (ppm))	0265	4.685	0.3900	5.563	0.7111	0.2343	16	-1.23	017.42	0
017	Boron (mg / kg (ppm))	0345	4.750	0.3000	5.563	0.7111	0.2343	16	-1.14	017.43	0
017	Boron (mg / kg (ppm))	0045	4.920	0.1400	5.563	0.7111	0.2343	16	-0.90	017.42	0
017	Boron (mg / kg (ppm))	0693	4.948	0.3070	5.563	0.7111	0.2343	16	-0.87	017.42	0
017	Boron (mg / kg (ppm))	0358	5.110	0.3200	5.563	0.7111	0.2343	16	-0.64	017.41	0
017	Boron (mg / kg (ppm))	0560	5.165	0.1300	5.563	0.7111	0.2343	16	-0.56	017.42	0
017	Boron (mg / kg (ppm))	2055	5.400	0.2000	5.563	0.7111	0.2343	16	-0.23	017.43	0
017	Boron (mg / kg (ppm))	0294	5.450	0.5000	5.563	0.7111	0.2343	16	-0.16	017.42	0
017	Boron (mg / kg (ppm))	0083	5.550	0.3000	5.563	0.7111	0.2343	16	-0.02	017.43	0
017	Boron (mg / kg (ppm))	0407	5.691	0.0022	5.563	0.7111	0.2343	16	0.18	017.41	0
017	Boron (mg / kg (ppm))	0918	5.720	0.0200	5.563	0.7111	0.2343	16	0.22	017.43	0
017	Boron (mg / kg (ppm))	0510	6.000	0.0000	5.563	0.7111	0.2343	16	0.61	017.43	0
017	Boron (mg / kg (ppm))	2053	6.150	0.7000	5.563	0.7111	0.2343	16	0.83	017.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
017	Boron (mg / kg (ppm))	0226	6.210	0.0200	5.563	0.7111	0.2343	16	0.91	017.41	0
017	Boron (mg / kg (ppm))	0049	6.820	0.4200	5.563	0.7111	0.2343	16	1.77	017.41	0
017	Boron (mg / kg (ppm))	0353	8.000	0.0000	5.563	0.7111	0.2343	16	3.43	017.43	0
017	Boron (mg / kg (ppm))	0229	4.430	1.300	5.563	0.7111	0.2343	16	-1.59	017.41	1
019	Calcium (%)	0873	1.050	0.0200	1.677	0.0852	0.0343	158	-7.36	019.44	0
019	Calcium (%)	0852	1.150	0.1000	1.677	0.0852	0.0343	158	-6.19	019.08	0
019	Calcium (%)	0889	1.175	0.0100	1.677	0.0852	0.0343	158	-5.89	019.99	0
019	Calcium (%)	2031	1.242	0.0477	1.677	0.0852	0.0343	158	-5.11	019.08	0
019	Calcium (%)	2129	1.355	0.0100	1.677	0.0852	0.0343	158	-3.78	019.42	0
019	Calcium (%)	2128	1.370	0.0030	1.677	0.0852	0.0343	158	-3.61	019.02	0
019	Calcium (%)	0609	1.420	0.0400	1.677	0.0852	0.0343	158	-3.02	019.32	0
019	Calcium (%)	2006	1.442	0.0380	1.677	0.0852	0.0343	158	-2.76	019.00	0
019	Calcium (%)	0695	1.502	0.0210	1.677	0.0852	0.0343	158	-2.06	019.41	0
019	Calcium (%)	0144	1.530	0.0600	1.677	0.0852	0.0343	158	-1.73	019.41	0
019	Calcium (%)	0676	1.530	0.0000	1.677	0.0852	0.0343	158	-1.73	019.99	0
019	Calcium (%)	0536	1.535	0.0100	1.677	0.0852	0.0343	158	-1.67	019.31	0
019	Calcium (%)	2198	1.535	0.0300	1.677	0.0852	0.0343	158	-1.67	019.31	0
019	Calcium (%)	0693	1.536	0.0090	1.677	0.0852	0.0343	158	-1.66	019.42	0
019	Calcium (%)	2018	1.540	0.0400	1.677	0.0852	0.0343	158	-1.61	019.31	0
019	Calcium (%)	0914	1.550	0.0036	1.677	0.0852	0.0343	158	-1.49	019.43	0
019	Calcium (%)	2006	1.560	0.0320	1.677	0.0852	0.0343	158	-1.37	019.09	0
019	Calcium (%)	0656	1.570	0.0400	1.677	0.0852	0.0343	158	-1.26	019.31	0
019	Calcium (%)	2220	1.570	0.0400	1.677	0.0852	0.0343	158	-1.26	019.44	0
019	Calcium (%)	0004	1.575	0.1500	1.677	0.0852	0.0343	158	-1.20	019.41	0
019	Calcium (%)	0098	1.575	0.0100	1.677	0.0852	0.0343	158	-1.20	019.41	0
019	Calcium (%)	0171	1.575	0.0100	1.677	0.0852	0.0343	158	-1.20	019.41	0
019	Calcium (%)	2055	1.577	0.0424	1.677	0.0852	0.0343	158	-1.18	019.43	0
019	Calcium (%)	2109	1.583	0.0540	1.677	0.0852	0.0343	158	-1.10	019.00	0
019	Calcium (%)	0689	1.595	0.0500	1.677	0.0852	0.0343	158	-0.96	019.08	0
019	Calcium (%)	2146	1.595	0.0500	1.677	0.0852	0.0343	158	-0.96	019.31	0
019	Calcium (%)	0278	1.600	0.0200	1.677	0.0852	0.0343	158	-0.90	019.42	0
019	Calcium (%)	2089	1.600	0.0000	1.677	0.0852	0.0343	158	-0.90	019.43	0
019	Calcium (%)	0037	1.602	0.0260	1.677	0.0852	0.0343	158	-0.88	019.42	0
019	Calcium (%)	2187	1.604	0.0084	1.677	0.0852	0.0343	158	-0.86	019.43	0
019	Calcium (%)	2115	1.605	0.0500	1.677	0.0852	0.0343	158	-0.85	019.31	0
019	Calcium (%)	0139	1.610	0.0000	1.677	0.0852	0.0343	158	-0.79	019.41	0
019	Calcium (%)	0205	1.610	0.0000	1.677	0.0852	0.0343	158	-0.79	019.42	0
019	Calcium (%)	2058	1.611	0.0140	1.677	0.0852	0.0343	158	-0.78	019.43	0
019	Calcium (%)	0590	1.615	0.0500	1.677	0.0852	0.0343	158	-0.73	019.99	0
019	Calcium (%)	0169	1.615	0.0100	1.677	0.0852	0.0343	158	-0.73	019.32	0
019	Calcium (%)	0202	1.620	0.0410	1.677	0.0852	0.0343	158	-0.68	019.43	0
019	Calcium (%)	0511	1.620	0.0400	1.677	0.0852	0.0343	158	-0.67	019.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	2181	1.625	0.0500	1.677	0.0852	0.0343	158	-0.61	019.00	0
019	Calcium (%)	0051	1.625	0.0100	1.677	0.0852	0.0343	158	-0.61	019.41	0
019	Calcium (%)	2012	1.625	0.0100	1.677	0.0852	0.0343	158	-0.61	019.41	0
019	Calcium (%)	2219	1.625	0.0300	1.677	0.0852	0.0343	158	-0.61	019.44	0
019	Calcium (%)	2146	1.625	0.0900	1.677	0.0852	0.0343	158	-0.61	019.51	0
019	Calcium (%)	0011	1.627	0.0810	1.677	0.0852	0.0343	158	-0.59	019.41	0
019	Calcium (%)	0674	1.630	0.0000	1.677	0.0852	0.0343	158	-0.55	019.31	0
019	Calcium (%)	0675	1.630	0.0200	1.677	0.0852	0.0343	158	-0.55	019.43	0
019	Calcium (%)	2252	1.630	0.0800	1.677	0.0852	0.0343	158	-0.55	019.44	0
019	Calcium (%)	0651	1.631	0.0720	1.677	0.0852	0.0343	158	-0.54	019.00	0
019	Calcium (%)	0154	1.633	0.1189	1.677	0.0852	0.0343	158	-0.52	019.52	0
019	Calcium (%)	0298	1.640	0.0200	1.677	0.0852	0.0343	158	-0.44	019.41	0
019	Calcium (%)	0065	1.640	0.0000	1.677	0.0852	0.0343	158	-0.44	019.44	0
019	Calcium (%)	0510	1.640	0.0400	1.677	0.0852	0.0343	158	-0.44	019.43	0
019	Calcium (%)	2022	1.645	0.0100	1.677	0.0852	0.0343	158	-0.38	019.31	0
019	Calcium (%)	0003	1.645	0.0900	1.677	0.0852	0.0343	158	-0.38	019.41	0
019	Calcium (%)	0345	1.645	0.0300	1.677	0.0852	0.0343	158	-0.38	019.43	0
019	Calcium (%)	2246	1.650	0.0000	1.677	0.0852	0.0343	158	-0.32	019.31	0
019	Calcium (%)	0010	1.650	0.0000	1.677	0.0852	0.0343	158	-0.32	019.43	0
019	Calcium (%)	0425	1.650	0.0200	1.677	0.0852	0.0343	158	-0.32	019.43	0
019	Calcium (%)	0572	1.650	0.0200	1.677	0.0852	0.0343	158	-0.32	019.53	0
019	Calcium (%)	0874	1.655	0.0100	1.677	0.0852	0.0343	158	-0.26	019.31	0
019	Calcium (%)	0884	1.655	0.0100	1.677	0.0852	0.0343	158	-0.26	019.31	0
019	Calcium (%)	0860	1.658	0.0000	1.677	0.0852	0.0343	158	-0.23	019.43	0
019	Calcium (%)	0142	1.658	0.0143	1.677	0.0852	0.0343	158	-0.22	019.31	0
019	Calcium (%)	0885	1.659	0.0070	1.677	0.0852	0.0343	158	-0.22	019.08	0
019	Calcium (%)	0848	1.660	0.0000	1.677	0.0852	0.0343	158	-0.20	019.41	0
019	Calcium (%)	0036	1.663	0.0033	1.677	0.0852	0.0343	158	-0.17	019.43	0
019	Calcium (%)	0164	1.665	0.0100	1.677	0.0852	0.0343	158	-0.14	019.41	0
019	Calcium (%)	0955	1.665	0.1100	1.677	0.0852	0.0343	158	-0.14	019.44	0
019	Calcium (%)	2114	1.668	0.0399	1.677	0.0852	0.0343	158	-0.11	019.44	0
019	Calcium (%)	0878	1.670	0.0400	1.677	0.0852	0.0343	158	-0.08	019.00	0
019	Calcium (%)	2246	1.670	0.0000	1.677	0.0852	0.0343	158	-0.08	019.00	0
019	Calcium (%)	0066	1.670	0.0800	1.677	0.0852	0.0343	158	-0.08	019.41	0
019	Calcium (%)	0893	1.675	0.0100	1.677	0.0852	0.0343	158	-0.02	019.00	0
019	Calcium (%)	2126	1.675	0.0300	1.677	0.0852	0.0343	158	-0.02	019.31	0
019	Calcium (%)	0619	1.675	0.0100	1.677	0.0852	0.0343	158	-0.02	019.41	0
019	Calcium (%)	0186	1.675	0.0300	1.677	0.0852	0.0343	158	-0.02	019.52	0
019	Calcium (%)	0083	1.676	0.0050	1.677	0.0852	0.0343	158	-0.02	019.43	0
019	Calcium (%)	0229	1.678	0.0380	1.677	0.0852	0.0343	158	0.01	019.41	0
019	Calcium (%)	0668	1.680	0.1200	1.677	0.0852	0.0343	158	0.03	019.43	0
019	Calcium (%)	0653	1.683	0.0390	1.677	0.0852	0.0343	158	0.06	019.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	0590	1.685	0.0900	1.677	0.0852	0.0343	158	0.09	019.08	0
019	Calcium (%)	0045	1.685	0.0500	1.677	0.0852	0.0343	158	0.09	019.42	0
019	Calcium (%)	0870	1.688	0.0900	1.677	0.0852	0.0343	158	0.13	019.42	0
019	Calcium (%)	2188	1.690	0.0200	1.677	0.0852	0.0343	158	0.15	019.31	0
019	Calcium (%)	0074	1.695	0.0100	1.677	0.0852	0.0343	158	0.21	019.41	0
019	Calcium (%)	0520	1.695	0.0700	1.677	0.0852	0.0343	158	0.21	019.41	0
019	Calcium (%)	0033	1.695	0.1300	1.677	0.0852	0.0343	158	0.21	019.43	0
019	Calcium (%)	0098	1.695	0.0500	1.677	0.0852	0.0343	158	0.21	019.43	0
019	Calcium (%)	0560	1.700	0.0305	1.677	0.0852	0.0343	158	0.26	019.42	0
019	Calcium (%)	2196	1.700	0.0000	1.677	0.0852	0.0343	158	0.27	019.31	0
019	Calcium (%)	2216	1.700	0.1200	1.677	0.0852	0.0343	158	0.27	019.35	0
019	Calcium (%)	0019	1.700	0.0200	1.677	0.0852	0.0343	158	0.27	019.41	0
019	Calcium (%)	0405	1.700	0.0000	1.677	0.0852	0.0343	158	0.27	019.41	0
019	Calcium (%)	0366	1.700	0.1000	1.677	0.0852	0.0343	158	0.27	019.42	0
019	Calcium (%)	2033	1.700	0.0000	1.677	0.0852	0.0343	158	0.27	019.43	0
019	Calcium (%)	0939	1.700	0.0400	1.677	0.0852	0.0343	158	0.27	019.53	0
019	Calcium (%)	0242	1.700	0.0800	1.677	0.0852	0.0343	158	0.27	019.99	0
019	Calcium (%)	0881	1.700	0.0060	1.677	0.0852	0.0343	158	0.27	019.00	0
019	Calcium (%)	0598	1.702	0.0410	1.677	0.0852	0.0343	158	0.29	019.41	0
019	Calcium (%)	0948	1.702	0.0180	1.677	0.0852	0.0343	158	0.29	019.31	0
019	Calcium (%)	0049	1.705	0.0500	1.677	0.0852	0.0343	158	0.33	019.41	0
019	Calcium (%)	0354	1.705	0.0100	1.677	0.0852	0.0343	158	0.33	019.41	0
019	Calcium (%)	0208	1.706	0.1500	1.677	0.0852	0.0343	158	0.34	019.41	0
019	Calcium (%)	0968	1.707	0.0050	1.677	0.0852	0.0343	158	0.35	019.43	0
019	Calcium (%)	0017	1.708	0.0080	1.677	0.0852	0.0343	158	0.36	019.43	0
019	Calcium (%)	2196	1.710	0.0000	1.677	0.0852	0.0343	158	0.39	019.08	0
019	Calcium (%)	2259	1.710	0.0000	1.677	0.0852	0.0343	158	0.39	019.42	0
019	Calcium (%)	0921	1.713	0.0320	1.677	0.0852	0.0343	158	0.42	019.31	0
019	Calcium (%)	2144	1.715	0.0100	1.677	0.0852	0.0343	158	0.44	019.00	0
019	Calcium (%)	0687	1.715	0.1300	1.677	0.0852	0.0343	158	0.44	019.31	0
019	Calcium (%)	0186	1.715	0.0100	1.677	0.0852	0.0343	158	0.44	019.42	0
019	Calcium (%)	2156	1.718	0.0010	1.677	0.0852	0.0343	158	0.47	019.44	0
019	Calcium (%)	0035	1.720	0.0000	1.677	0.0852	0.0343	158	0.50	019.43	0
019	Calcium (%)	0089	1.720	0.0000	1.677	0.0852	0.0343	158	0.50	019.43	0
019	Calcium (%)	0407	1.725	0.0049	1.677	0.0852	0.0343	158	0.56	019.41	0
019	Calcium (%)	2193	1.725	0.0300	1.677	0.0852	0.0343	158	0.56	019.08	0
019	Calcium (%)	2259	1.725	0.0100	1.677	0.0852	0.0343	158	0.56	019.41	0
019	Calcium (%)	2188	1.730	0.0000	1.677	0.0852	0.0343	158	0.62	019.08	0
019	Calcium (%)	0878	1.730	0.0200	1.677	0.0852	0.0343	158	0.62	019.41	0
019	Calcium (%)	2009	1.732	0.0028	1.677	0.0852	0.0343	158	0.64	019.08	0
019	Calcium (%)	0512	1.733	0.0540	1.677	0.0852	0.0343	158	0.66	019.43	0
019	Calcium (%)	0683	1.735	0.0100	1.677	0.0852	0.0343	158	0.68	019.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	0610	1.735	0.0500	1.677	0.0852	0.0343	158	0.68	019.43	0
019	Calcium (%)	0726	1.739	0.0290	1.677	0.0852	0.0343	158	0.72	019.42	0
019	Calcium (%)	2248	1.740	0.0160	1.677	0.0852	0.0343	158	0.74	019.31	0
019	Calcium (%)	0964	1.742	0.0702	1.677	0.0852	0.0343	158	0.76	019.43	0
019	Calcium (%)	0868	1.745	0.0300	1.677	0.0852	0.0343	158	0.80	019.31	0
019	Calcium (%)	0047	1.745	0.1500	1.677	0.0852	0.0343	158	0.80	019.52	0
019	Calcium (%)	0208	1.746	0.0340	1.677	0.0852	0.0343	158	0.81	019.31	0
019	Calcium (%)	0910	1.750	0.0000	1.677	0.0852	0.0343	158	0.86	019.41	0
019	Calcium (%)	2207	1.755	0.0300	1.677	0.0852	0.0343	158	0.91	019.42	0
019	Calcium (%)	2192	1.755	0.0900	1.677	0.0852	0.0343	158	0.91	019.43	0
019	Calcium (%)	0263	1.764	0.0058	1.677	0.0852	0.0343	158	1.02	019.42	0
019	Calcium (%)	2190	1.765	0.0453	1.677	0.0852	0.0343	158	1.03	019.08	0
019	Calcium (%)	0505	1.765	0.0300	1.677	0.0852	0.0343	158	1.03	019.43	0
019	Calcium (%)	0294	1.770	0.0200	1.677	0.0852	0.0343	158	1.09	019.42	0
019	Calcium (%)	2115	1.775	0.0900	1.677	0.0852	0.0343	158	1.15	019.00	0
019	Calcium (%)	0849	1.779	0.0180	1.677	0.0852	0.0343	158	1.20	019.08	0
019	Calcium (%)	0001	1.780	0.0160	1.677	0.0852	0.0343	158	1.21	019.31	0
019	Calcium (%)	0226	1.780	0.0000	1.677	0.0852	0.0343	158	1.21	019.41	0
019	Calcium (%)	0035	1.780	0.0000	1.677	0.0852	0.0343	158	1.21	019.42	0
019	Calcium (%)	0682	1.780	0.0000	1.677	0.0852	0.0343	158	1.21	019.43	0
019	Calcium (%)	0918	1.780	0.0200	1.677	0.0852	0.0343	158	1.21	019.43	0
019	Calcium (%)	0028	1.790	0.0200	1.677	0.0852	0.0343	158	1.33	019.43	0
019	Calcium (%)	0036	1.794	0.0332	1.677	0.0852	0.0343	158	1.38	019.03	0
019	Calcium (%)	0897	1.795	0.0100	1.677	0.0852	0.0343	158	1.38	019.00	0
019	Calcium (%)	0358	1.795	0.0500	1.677	0.0852	0.0343	158	1.38	019.41	0
019	Calcium (%)	0187	1.795	0.0100	1.677	0.0852	0.0343	158	1.38	019.42	0
019	Calcium (%)	2222	1.800	0.0200	1.677	0.0852	0.0343	158	1.44	019.31	0
019	Calcium (%)	0042	1.810	0.1000	1.677	0.0852	0.0343	158	1.56	019.42	0
019	Calcium (%)	0529	1.820	0.0000	1.677	0.0852	0.0343	158	1.68	019.31	0
019	Calcium (%)	0108	1.835	0.0300	1.677	0.0852	0.0343	158	1.85	019.31	0
019	Calcium (%)	2124	1.845	0.0100	1.677	0.0852	0.0343	158	1.97	019.43	0
019	Calcium (%)	0190	1.865	0.0500	1.677	0.0852	0.0343	158	2.21	019.42	0
019	Calcium (%)	0100	1.865	0.0500	1.677	0.0852	0.0343	158	2.21	019.99	0
019	Calcium (%)	0353	1.870	0.0400	1.677	0.0852	0.0343	158	2.26	019.43	0
019	Calcium (%)	0032	1.950	0.1400	1.677	0.0852	0.0343	158	3.20	019.43	0
019	Calcium (%)	0357	1.957	0.0253	1.677	0.0852	0.0343	158	3.28	019.42	0
019	Calcium (%)	0038	1.511	0.2950	1.677	0.0852	0.0343	158	-1.96	019.43	1
019	Calcium (%)	0265	1.670	0.2200	1.677	0.0852	0.0343	158	-0.08	019.42	1
019	Calcium (%)	2053	1.670	0.2800	1.677	0.0852	0.0343	158	-0.08	019.42	1
019	Calcium (%)	0613	1.705	0.2100	1.677	0.0852	0.0343	158	0.33	019.42	1
021	Cobalt (mg / kg (ppm))	0910	0.2400	0.0200	0.8995	0.4638	0.1094	11	-1.42	021.52	0
021	Cobalt (mg / kg (ppm))	0918	0.3144	0.0208	0.8995	0.4638	0.1094	11	-1.26	021.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
021	Cobalt (mg / kg (ppm))	0186	0.4650	0.0180	0.8995	0.4638	0.1094	11	-0.94	021.52	0
021	Cobalt (mg / kg (ppm))	0345	0.8250	0.0300	0.8995	0.4638	0.1094	11	-0.16	021.43	0
021	Cobalt (mg / kg (ppm))	0038	0.8920	0.3060	0.8995	0.4638	0.1094	11	-0.02	021.43	0
021	Cobalt (mg / kg (ppm))	0171	0.9650	0.0300	0.8995	0.4638	0.1094	11	0.14	021.41	0
021	Cobalt (mg / kg (ppm))	0510	1.060	0.0000	0.8995	0.4638	0.1094	11	0.35	021.43	0
021	Cobalt (mg / kg (ppm))	0169	1.175	0.0300	0.8995	0.4638	0.1094	11	0.59	021.43	0
021	Cobalt (mg / kg (ppm))	0164	1.200	0.2000	0.8995	0.4638	0.1094	11	0.65	021.31	0
021	Cobalt (mg / kg (ppm))	0407	1.231	0.1188	0.8995	0.4638	0.1094	11	0.71	021.41	0
021	Cobalt (mg / kg (ppm))	0353	1.528	0.4302	0.8995	0.4638	0.1094	11	1.35	021.43	0
021	Cobalt (mg / kg (ppm))	0693	0.6600	0.8860	0.8995	0.4638	0.1094	11	-0.52	021.42	1
021	Cobalt (mg / kg (ppm))	2053	< 0.05		0.8995	0.4638	0.1094	11		021.42	5
021	Cobalt (mg / kg (ppm))	0560	< 1		0.8995	0.4638	0.1094	11		021.42	5
022	Copper (mg / kg (ppm))	0609	0.0350	0.0100	16.18	1.719	0.7171	95	-9.40	022.32	0
022	Copper (mg / kg (ppm))	2188	8.615	0.0300	16.18	1.719	0.7171	95	-4.40	022.31	0
022	Copper (mg / kg (ppm))	2246	10.36	0.7177	16.18	1.719	0.7171	95	-3.39	022.34	0
022	Copper (mg / kg (ppm))	0914	11.62	3.204	16.18	1.719	0.7171	95	-2.65	022.53	0
022	Copper (mg / kg (ppm))	0848	12.01	0.2400	16.18	1.719	0.7171	95	-2.43	022.41	0
022	Copper (mg / kg (ppm))	0505	12.50	1.920	16.18	1.719	0.7171	95	-2.14	022.43	0
022	Copper (mg / kg (ppm))	0868	12.57	0.0800	16.18	1.719	0.7171	95	-2.10	022.31	0
022	Copper (mg / kg (ppm))	0596	13.22	0.1400	16.18	1.719	0.7171	95	-1.72	022.41	0
022	Copper (mg / kg (ppm))	0536	13.40	0.1000	16.18	1.719	0.7171	95	-1.62	022.31	0
022	Copper (mg / kg (ppm))	2114	13.53	0.1757	16.18	1.719	0.7171	95	-1.54	022.44	0
022	Copper (mg / kg (ppm))	0619	13.55	1.900	16.18	1.719	0.7171	95	-1.53	022.41	0
022	Copper (mg / kg (ppm))	0045	14.40	0.2000	16.18	1.719	0.7171	95	-1.04	022.42	0
022	Copper (mg / kg (ppm))	0520	14.50	1.000	16.18	1.719	0.7171	95	-0.98	022.41	0
022	Copper (mg / kg (ppm))	0205	14.50	2.600	16.18	1.719	0.7171	95	-0.98	022.42	0
022	Copper (mg / kg (ppm))	0939	14.51	0.4300	16.18	1.719	0.7171	95	-0.98	022.53	0
022	Copper (mg / kg (ppm))	0051	14.64	0.4900	16.18	1.719	0.7171	95	-0.90	022.41	0
022	Copper (mg / kg (ppm))	0171	14.68	0.3800	16.18	1.719	0.7171	95	-0.87	022.41	0
022	Copper (mg / kg (ppm))	0668	14.70	0.6000	16.18	1.719	0.7171	95	-0.86	022.43	0
022	Copper (mg / kg (ppm))	0948	14.73	0.0600	16.18	1.719	0.7171	95	-0.84	022.33	0
022	Copper (mg / kg (ppm))	0407	14.75	0.3595	16.18	1.719	0.7171	95	-0.83	022.41	0
022	Copper (mg / kg (ppm))	0955	14.85	0.7000	16.18	1.719	0.7171	95	-0.78	022.41	0
022	Copper (mg / kg (ppm))	0910	15.00	0.0000	16.18	1.719	0.7171	95	-0.69	022.41	0
022	Copper (mg / kg (ppm))	0345	15.00	0.2000	16.18	1.719	0.7171	95	-0.69	022.43	0
022	Copper (mg / kg (ppm))	0510	15.00	0.0000	16.18	1.719	0.7171	95	-0.69	022.43	0
022	Copper (mg / kg (ppm))	0049	15.01	1.770	16.18	1.719	0.7171	95	-0.68	022.41	0
022	Copper (mg / kg (ppm))	0229	15.24	0.4700	16.18	1.719	0.7171	95	-0.55	022.41	0
022	Copper (mg / kg (ppm))	0693	15.32	0.7620	16.18	1.719	0.7171	95	-0.50	022.42	0
022	Copper (mg / kg (ppm))	0169	15.35	0.3000	16.18	1.719	0.7171	95	-0.48	022.43	0
022	Copper (mg / kg (ppm))	0572	15.35	0.3000	16.18	1.719	0.7171	95	-0.48	022.53	0
022	Copper (mg / kg (ppm))	0682	15.40	0.0000	16.18	1.719	0.7171	95	-0.46	022.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
022	Copper (mg / kg (ppm))	0074	15.50	1.000	16.18	1.719	0.7171	95	-0.40	022.41	0
022	Copper (mg / kg (ppm))	0366	15.50	1.000	16.18	1.719	0.7171	95	-0.40	022.42	0
022	Copper (mg / kg (ppm))	0100	15.50	1.000	16.18	1.719	0.7171	95	-0.40	022.99	0
022	Copper (mg / kg (ppm))	0278	15.55	0.3000	16.18	1.719	0.7171	95	-0.37	022.42	0
022	Copper (mg / kg (ppm))	2055	15.75	0.5000	16.18	1.719	0.7171	95	-0.25	022.43	0
022	Copper (mg / kg (ppm))	0033	15.80	0.6000	16.18	1.719	0.7171	95	-0.22	022.43	0
022	Copper (mg / kg (ppm))	0038	15.80	0.6000	16.18	1.719	0.7171	95	-0.22	022.43	0
022	Copper (mg / kg (ppm))	0613	15.95	0.9000	16.18	1.719	0.7171	95	-0.14	022.42	0
022	Copper (mg / kg (ppm))	0208	15.96	0.2700	16.18	1.719	0.7171	95	-0.13	022.41	0
022	Copper (mg / kg (ppm))	0164	16.00	0.0000	16.18	1.719	0.7171	95	-0.11	022.41	0
022	Copper (mg / kg (ppm))	0511	16.00	0.0000	16.18	1.719	0.7171	95	-0.11	022.41	0
022	Copper (mg / kg (ppm))	0968	16.00	0.0000	16.18	1.719	0.7171	95	-0.11	022.43	0
022	Copper (mg / kg (ppm))	0590	16.00	0.6000	16.18	1.719	0.7171	95	-0.11	022.99	0
022	Copper (mg / kg (ppm))	0037	16.05	0.3000	16.18	1.719	0.7171	95	-0.08	022.42	0
022	Copper (mg / kg (ppm))	0047	16.05	0.5800	16.18	1.719	0.7171	95	-0.08	022.52	0
022	Copper (mg / kg (ppm))	2089	16.17	0.8100	16.18	1.719	0.7171	95	-0.01	022.43	0
022	Copper (mg / kg (ppm))	0870	16.19	0.3800	16.18	1.719	0.7171	95	0.00	022.42	0
022	Copper (mg / kg (ppm))	0689	16.30	0.2000	16.18	1.719	0.7171	95	0.07	022.31	0
022	Copper (mg / kg (ppm))	0208	16.35	3.100	16.18	1.719	0.7171	95	0.10	022.31	0
022	Copper (mg / kg (ppm))	2115	16.39	0.2300	16.18	1.719	0.7171	95	0.12	022.31	0
022	Copper (mg / kg (ppm))	2207	16.50	1.000	16.18	1.719	0.7171	95	0.18	022.42	0
022	Copper (mg / kg (ppm))	0035	16.50	1.000	16.18	1.719	0.7171	95	0.18	022.43	0
022	Copper (mg / kg (ppm))	0083	16.50	1.000	16.18	1.719	0.7171	95	0.18	022.43	0
022	Copper (mg / kg (ppm))	0186	16.50	1.000	16.18	1.719	0.7171	95	0.18	022.52	0
022	Copper (mg / kg (ppm))	0242	16.50	1.000	16.18	1.719	0.7171	95	0.18	022.99	0
022	Copper (mg / kg (ppm))	2129	16.50	0.3200	16.18	1.719	0.7171	95	0.18	022.99	0
022	Copper (mg / kg (ppm))	0964	16.52	0.5827	16.18	1.719	0.7171	95	0.20	022.43	0
022	Copper (mg / kg (ppm))	0263	16.57	0.1470	16.18	1.719	0.7171	95	0.22	022.42	0
022	Copper (mg / kg (ppm))	0019	16.58	1.150	16.18	1.719	0.7171	95	0.23	022.41	0
022	Copper (mg / kg (ppm))	0098	16.62	0.2400	16.18	1.719	0.7171	95	0.25	022.41	0
022	Copper (mg / kg (ppm))	0675	16.63	0.3300	16.18	1.719	0.7171	95	0.26	022.43	0
022	Copper (mg / kg (ppm))	0560	16.65	0.1000	16.18	1.719	0.7171	95	0.27	022.42	0
022	Copper (mg / kg (ppm))	0354	16.77	0.0100	16.18	1.719	0.7171	95	0.34	022.41	0
022	Copper (mg / kg (ppm))	0674	16.80	0.0000	16.18	1.719	0.7171	95	0.36	022.31	0
022	Copper (mg / kg (ppm))	0098	16.82	0.3200	16.18	1.719	0.7171	95	0.37	022.43	0
022	Copper (mg / kg (ppm))	0226	16.91	1.190	16.18	1.719	0.7171	95	0.42	022.41	0
022	Copper (mg / kg (ppm))	0035	16.91	0.1800	16.18	1.719	0.7171	95	0.42	022.42	0
022	Copper (mg / kg (ppm))	0425	16.98	0.0600	16.18	1.719	0.7171	95	0.46	022.43	0
022	Copper (mg / kg (ppm))	0610	17.00	0.0000	16.18	1.719	0.7171	95	0.48	022.43	0
022	Copper (mg / kg (ppm))	0358	17.03	0.7000	16.18	1.719	0.7171	95	0.49	022.41	0
022	Copper (mg / kg (ppm))	0010	17.10	0.4000	16.18	1.719	0.7171	95	0.53	022.33	0
022	Copper (mg / kg (ppm))	0042	17.15	1.700	16.18	1.719	0.7171	95	0.56	022.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
022	Copper (mg / kg (ppm))	0512	17.16	0.2200	16.18	1.719	0.7171	95	0.57	022.43	0
022	Copper (mg / kg (ppm))	2053	17.25	0.7000	16.18	1.719	0.7171	95	0.62	022.42	0
022	Copper (mg / kg (ppm))	0357	17.37	0.6930	16.18	1.719	0.7171	95	0.69	022.42	0
022	Copper (mg / kg (ppm))	0190	17.48	0.7100	16.18	1.719	0.7171	95	0.75	022.42	0
022	Copper (mg / kg (ppm))	0726	17.48	0.0100	16.18	1.719	0.7171	95	0.75	022.42	0
022	Copper (mg / kg (ppm))	0186	17.50	1.000	16.18	1.719	0.7171	95	0.77	022.42	0
022	Copper (mg / kg (ppm))	0294	17.68	0.2400	16.18	1.719	0.7171	95	0.87	022.42	0
022	Copper (mg / kg (ppm))	2124	17.69	0.1900	16.18	1.719	0.7171	95	0.87	022.43	0
022	Copper (mg / kg (ppm))	0598	17.95	0.5300	16.18	1.719	0.7171	95	1.03	022.41	0
022	Copper (mg / kg (ppm))	2033	17.95	0.3000	16.18	1.719	0.7171	95	1.03	022.43	0
022	Copper (mg / kg (ppm))	0202	18.20	1.278	16.18	1.719	0.7171	95	1.17	022.43	0
022	Copper (mg / kg (ppm))	2196	18.40	0.0000	16.18	1.719	0.7171	95	1.29	022.31	0
022	Copper (mg / kg (ppm))	0353	18.50	3.000	16.18	1.719	0.7171	95	1.35	022.43	0
022	Copper (mg / kg (ppm))	0066	18.52	1.910	16.18	1.719	0.7171	95	1.36	022.44	0
022	Copper (mg / kg (ppm))	0878	19.00	4.000	16.18	1.719	0.7171	95	1.64	022.41	0
022	Copper (mg / kg (ppm))	0265	19.70	1.200	16.18	1.719	0.7171	95	2.05	022.42	0
022	Copper (mg / kg (ppm))	0695	19.93	0.4660	16.18	1.719	0.7171	95	2.18	022.41	0
022	Copper (mg / kg (ppm))	0004	20.95	1.110	16.18	1.719	0.7171	95	2.77	022.41	0
022	Copper (mg / kg (ppm))	0918	20.98	4.340	16.18	1.719	0.7171	95	2.79	022.43	0
022	Copper (mg / kg (ppm))	2012	22.41	0.5900	16.18	1.719	0.7171	95	3.62	022.41	0
022	Copper (mg / kg (ppm))	0187	22.85	0.7000	16.18	1.719	0.7171	95	3.88	022.42	0
022	Copper (mg / kg (ppm))	0003	25.50	1.000	16.18	1.719	0.7171	95	5.42	022.41	0
022	Copper (mg / kg (ppm))	0405	27.46	0.0000	16.18	1.719	0.7171	95	6.56	022.41	0
022	Copper (mg / kg (ppm))	2192	19.62	6.570	16.18	1.719	0.7171	95	2.00	022.43	1
022	Copper (mg / kg (ppm))	0529	23.85	4.900	16.18	1.719	0.7171	95	4.46	022.31	1
022	Copper (mg / kg (ppm))	0011	58.79	4.829	16.18	1.719	0.7171	95	24.79	022.41	2
022	Copper (mg / kg (ppm))	0656	< 20.2		16.18	1.719	0.7171	95		022.31	5
025	Iron (mg / kg (ppm))	0609	0.3050	0.0100	297.9	27.91	11.05	94	-10.66	025.32	0
025	Iron (mg / kg (ppm))	0914	28.56	0.6000	297.9	27.91	11.05	94	-9.65	025.43	0
025	Iron (mg / kg (ppm))	0921	30.59	2.360	297.9	27.91	11.05	94	-9.58	025.31	0
025	Iron (mg / kg (ppm))	0278	131.0	10.00	297.9	27.91	11.05	94	-5.98	025.42	0
025	Iron (mg / kg (ppm))	2053	182.0	28.00	297.9	27.91	11.05	94	-4.15	025.42	0
025	Iron (mg / kg (ppm))	0689	197.9	10.00	297.9	27.91	11.05	94	-3.58	025.31	0
025	Iron (mg / kg (ppm))	2246	240.1	8.326	297.9	27.91	11.05	94	-2.07	025.31	0
025	Iron (mg / kg (ppm))	0047	240.7	4.200	297.9	27.91	11.05	94	-2.05	025.52	0
025	Iron (mg / kg (ppm))	2114	241.2	5.276	297.9	27.91	11.05	94	-2.03	025.41	0
025	Iron (mg / kg (ppm))	0726	242.1	1.600	297.9	27.91	11.05	94	-2.00	025.42	0
025	Iron (mg / kg (ppm))	0955	248.0	2.000	297.9	27.91	11.05	94	-1.79	025.41	0
025	Iron (mg / kg (ppm))	0653	257.1	0.5000	297.9	27.91	11.05	94	-1.46	025.41	0
025	Iron (mg / kg (ppm))	0598	257.2	5.300	297.9	27.91	11.05	94	-1.46	025.41	0
025	Iron (mg / kg (ppm))	0294	267.1	19.48	297.9	27.91	11.05	94	-1.11	025.42	0
025	Iron (mg / kg (ppm))	0874	268.0	6.000	297.9	27.91	11.05	94	-1.07	025.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
025	Iron (mg / kg (ppm))	2012	274.0	4.170	297.9	27.91	11.05	94	-0.86	025.41	0
025	Iron (mg / kg (ppm))	0668	274.5	13.00	297.9	27.91	11.05	94	-0.84	025.43	0
025	Iron (mg / kg (ppm))	0001	275.5	8.800	297.9	27.91	11.05	94	-0.80	025.31	0
025	Iron (mg / kg (ppm))	0407	276.0	3.446	297.9	27.91	11.05	94	-0.79	025.41	0
025	Iron (mg / kg (ppm))	0510	276.5	1.000	297.9	27.91	11.05	94	-0.77	025.43	0
025	Iron (mg / kg (ppm))	0619	278.5	9.000	297.9	27.91	11.05	94	-0.70	025.41	0
025	Iron (mg / kg (ppm))	0695	279.4	2.529	297.9	27.91	11.05	94	-0.66	025.41	0
025	Iron (mg / kg (ppm))	0345	279.6	2.900	297.9	27.91	11.05	94	-0.66	025.43	0
025	Iron (mg / kg (ppm))	0870	280.1	12.30	297.9	27.91	11.05	94	-0.64	025.42	0
025	Iron (mg / kg (ppm))	0354	281.6	0.3800	297.9	27.91	11.05	94	-0.59	025.41	0
025	Iron (mg / kg (ppm))	0037	282.0	1.000	297.9	27.91	11.05	94	-0.57	025.42	0
025	Iron (mg / kg (ppm))	0226	282.1	9.360	297.9	27.91	11.05	94	-0.57	025.41	0
025	Iron (mg / kg (ppm))	0190	282.3	10.68	297.9	27.91	11.05	94	-0.56	025.42	0
025	Iron (mg / kg (ppm))	2192	283.5	3.980	297.9	27.91	11.05	94	-0.52	025.43	0
025	Iron (mg / kg (ppm))	0011	284.8	7.605	297.9	27.91	11.05	94	-0.47	025.41	0
025	Iron (mg / kg (ppm))	0878	285.5	11.00	297.9	27.91	11.05	94	-0.45	025.41	0
025	Iron (mg / kg (ppm))	0693	287.2	15.78	297.9	27.91	11.05	94	-0.38	025.42	0
025	Iron (mg / kg (ppm))	0164	289.5	5.000	297.9	27.91	11.05	94	-0.30	025.41	0
025	Iron (mg / kg (ppm))	0265	289.5	5.000	297.9	27.91	11.05	94	-0.30	025.42	0
025	Iron (mg / kg (ppm))	0045	292.0	0.0000	297.9	27.91	11.05	94	-0.21	025.42	0
025	Iron (mg / kg (ppm))	0033	293.5	17.00	297.9	27.91	11.05	94	-0.16	025.43	0
025	Iron (mg / kg (ppm))	0003	294.5	13.00	297.9	27.91	11.05	94	-0.12	025.41	0
025	Iron (mg / kg (ppm))	2124	295.1	35.30	297.9	27.91	11.05	94	-0.10	025.43	0
025	Iron (mg / kg (ppm))	2022	296.0	8.000	297.9	27.91	11.05	94	-0.07	025.31	0
025	Iron (mg / kg (ppm))	0066	297.0	10.14	297.9	27.91	11.05	94	-0.04	025.41	0
025	Iron (mg / kg (ppm))	0868	297.3	4.110	297.9	27.91	11.05	94	-0.02	025.31	0
025	Iron (mg / kg (ppm))	0520	297.5	1.000	297.9	27.91	11.05	94	-0.02	025.41	0
025	Iron (mg / kg (ppm))	0083	297.5	3.000	297.9	27.91	11.05	94	-0.02	025.43	0
025	Iron (mg / kg (ppm))	0529	298.1	15.40	297.9	27.91	11.05	94	0.01	025.31	0
025	Iron (mg / kg (ppm))	0613	298.7	27.90	297.9	27.91	11.05	94	0.03	025.42	0
025	Iron (mg / kg (ppm))	2146	299.1	17.86	297.9	27.91	11.05	94	0.04	025.31	0
025	Iron (mg / kg (ppm))	2207	299.5	15.00	297.9	27.91	11.05	94	0.06	025.42	0
025	Iron (mg / kg (ppm))	0425	299.6	7.800	297.9	27.91	11.05	94	0.06	025.43	0
025	Iron (mg / kg (ppm))	0049	300.2	9.670	297.9	27.91	11.05	94	0.08	025.41	0
025	Iron (mg / kg (ppm))	0169	302.0	6.000	297.9	27.91	11.05	94	0.15	025.43	0
025	Iron (mg / kg (ppm))	0051	302.5	1.000	297.9	27.91	11.05	94	0.16	025.41	0
025	Iron (mg / kg (ppm))	0560	303.0	28.10	297.9	27.91	11.05	94	0.18	025.42	0
025	Iron (mg / kg (ppm))	0511	303.0	10.00	297.9	27.91	11.05	94	0.18	025.41	0
025	Iron (mg / kg (ppm))	0910	303.0	12.00	297.9	27.91	11.05	94	0.18	025.41	0
025	Iron (mg / kg (ppm))	2196	303.4	0.0000	297.9	27.91	11.05	94	0.20	025.31	0
025	Iron (mg / kg (ppm))	0918	303.8	1.330	297.9	27.91	11.05	94	0.21	025.43	0
025	Iron (mg / kg (ppm))	0017	304.6	52.80	297.9	27.91	11.05	94	0.24	025.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
025	Iron (mg / kg (ppm))	2188	304.7	4.400	297.9	27.91	11.05	94	0.24	025.31	0
025	Iron (mg / kg (ppm))	0968	305.0	2.000	297.9	27.91	11.05	94	0.25	025.43	0
025	Iron (mg / kg (ppm))	0100	306.5	3.000	297.9	27.91	11.05	94	0.31	025.99	0
025	Iron (mg / kg (ppm))	0098	307.1	9.400	297.9	27.91	11.05	94	0.33	025.41	0
025	Iron (mg / kg (ppm))	0353	308.0	20.00	297.9	27.91	11.05	94	0.36	025.43	0
025	Iron (mg / kg (ppm))	0074	309.0	2.000	297.9	27.91	11.05	94	0.40	025.41	0
025	Iron (mg / kg (ppm))	0171	313.8	2.800	297.9	27.91	11.05	94	0.57	025.41	0
025	Iron (mg / kg (ppm))	0505	314.1	15.80	297.9	27.91	11.05	94	0.58	025.43	0
025	Iron (mg / kg (ppm))	0366	314.5	9.000	297.9	27.91	11.05	94	0.59	025.42	0
025	Iron (mg / kg (ppm))	2033	315.0	10.00	297.9	27.91	11.05	94	0.61	025.43	0
025	Iron (mg / kg (ppm))	0242	317.0	6.000	297.9	27.91	11.05	94	0.68	025.99	0
025	Iron (mg / kg (ppm))	0229	317.3	7.900	297.9	27.91	11.05	94	0.69	025.41	0
025	Iron (mg / kg (ppm))	0208	318.0	10.00	297.9	27.91	11.05	94	0.72	025.31	0
025	Iron (mg / kg (ppm))	0004	318.4	23.83	297.9	27.91	11.05	94	0.73	025.41	0
025	Iron (mg / kg (ppm))	0675	318.7	13.41	297.9	27.91	11.05	94	0.74	025.43	0
025	Iron (mg / kg (ppm))	0610	319.5	9.000	297.9	27.91	11.05	94	0.77	025.43	0
025	Iron (mg / kg (ppm))	0187	319.8	1.100	297.9	27.91	11.05	94	0.78	025.42	0
025	Iron (mg / kg (ppm))	0939	320.5	5.000	297.9	27.91	11.05	94	0.81	025.53	0
025	Iron (mg / kg (ppm))	0682	320.5	0.0000	297.9	27.91	11.05	94	0.81	025.43	0
025	Iron (mg / kg (ppm))	0019	321.0	29.10	297.9	27.91	11.05	94	0.82	025.41	0
025	Iron (mg / kg (ppm))	0263	321.4	2.170	297.9	27.91	11.05	94	0.84	025.42	0
025	Iron (mg / kg (ppm))	0512	321.7	11.90	297.9	27.91	11.05	94	0.85	025.43	0
025	Iron (mg / kg (ppm))	2144	321.7	2.990	297.9	27.91	11.05	94	0.85	025.31	0
025	Iron (mg / kg (ppm))	0035	322.5	7.000	297.9	27.91	11.05	94	0.88	025.42	0
025	Iron (mg / kg (ppm))	2089	322.9	33.74	297.9	27.91	11.05	94	0.89	025.43	0
025	Iron (mg / kg (ppm))	0098	324.0	11.90	297.9	27.91	11.05	94	0.93	025.43	0
025	Iron (mg / kg (ppm))	0208	324.8	19.10	297.9	27.91	11.05	94	0.96	025.41	0
025	Iron (mg / kg (ppm))	0358	325.8	23.86	297.9	27.91	11.05	94	1.00	025.41	0
025	Iron (mg / kg (ppm))	2146	328.9	5.300	297.9	27.91	11.05	94	1.11	025.51	0
025	Iron (mg / kg (ppm))	2115	336.2	3.530	297.9	27.91	11.05	94	1.37	025.31	0
025	Iron (mg / kg (ppm))	0035	345.5	21.00	297.9	27.91	11.05	94	1.70	025.43	0
025	Iron (mg / kg (ppm))	0042	347.5	77.00	297.9	27.91	11.05	94	1.78	025.42	0
025	Iron (mg / kg (ppm))	0032	349.5	19.00	297.9	27.91	11.05	94	1.85	025.43	0
025	Iron (mg / kg (ppm))	0038	356.0	12.00	297.9	27.91	11.05	94	2.08	025.43	0
025	Iron (mg / kg (ppm))	2129	356.4	62.10	297.9	27.91	11.05	94	2.09	025.99	0
025	Iron (mg / kg (ppm))	0405	361.6	0.0000	297.9	27.91	11.05	94	2.28	025.41	0
025	Iron (mg / kg (ppm))	0656	373.6	15.66	297.9	27.91	11.05	94	2.71	025.31	0
025	Iron (mg / kg (ppm))	0536	173.6	340.3	297.9	27.91	11.05	94	-4.45	025.31	1
027	Magnesium (%)	0011	0.0205	0.0010	0.2112	0.0124	0.0065	114	-15.33	027.41	0
027	Magnesium (%)	0873	0.0535	0.0030	0.2112	0.0124	0.0065	114	-12.68	027.44	0
027	Magnesium (%)	0889	0.1400	0.0000	0.2112	0.0124	0.0065	114	-5.72	027.99	0
027	Magnesium (%)	0689	0.1700	0.0000	0.2112	0.0124	0.0065	114	-3.31	027.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
027	Magnesium (%)	0066	0.1800	0.0100	0.2112	0.0124	0.0065	114	-2.51	027.41	0
027	Magnesium (%)	0407	0.1885	0.0058	0.2112	0.0124	0.0065	114	-1.82	027.41	0
027	Magnesium (%)	0154	0.1887	0.0171	0.2112	0.0124	0.0065	114	-1.81	027.52	0
027	Magnesium (%)	0668	0.1890	0.0160	0.2112	0.0124	0.0065	114	-1.78	027.43	0
027	Magnesium (%)	0609	0.1900	0.0200	0.2112	0.0124	0.0065	114	-1.70	027.32	0
027	Magnesium (%)	0596	0.1900	0.0200	0.2112	0.0124	0.0065	114	-1.70	027.41	0
027	Magnesium (%)	2055	0.1905	0.0021	0.2112	0.0124	0.0065	114	-1.67	027.43	0
027	Magnesium (%)	0345	0.1910	0.0020	0.2112	0.0124	0.0065	114	-1.62	027.43	0
027	Magnesium (%)	2129	0.1920	0.0057	0.2112	0.0124	0.0065	114	-1.54	027.42	0
027	Magnesium (%)	0870	0.1938	0.0105	0.2112	0.0124	0.0065	114	-1.40	027.42	0
027	Magnesium (%)	0004	0.1950	0.0100	0.2112	0.0124	0.0065	114	-1.30	027.41	0
027	Magnesium (%)	0229	0.1965	0.0050	0.2112	0.0124	0.0065	114	-1.18	027.41	0
027	Magnesium (%)	0529	0.1980	0.0000	0.2112	0.0124	0.0065	114	-1.06	027.31	0
027	Magnesium (%)	0202	0.1980	0.0080	0.2112	0.0124	0.0065	114	-1.06	027.43	0
027	Magnesium (%)	0047	0.1980	0.0120	0.2112	0.0124	0.0065	114	-1.06	027.52	0
027	Magnesium (%)	0037	0.1995	0.0010	0.2112	0.0124	0.0065	114	-0.94	027.42	0
027	Magnesium (%)	0511	0.2000	0.0000	0.2112	0.0124	0.0065	114	-0.90	027.41	0
027	Magnesium (%)	0510	0.2000	0.0000	0.2112	0.0124	0.0065	114	-0.90	027.43	0
027	Magnesium (%)	2058	0.2000	0.0020	0.2112	0.0124	0.0065	114	-0.90	027.43	0
027	Magnesium (%)	2089	0.2000	0.0000	0.2112	0.0124	0.0065	114	-0.90	027.43	0
027	Magnesium (%)	0948	0.2005	0.0010	0.2112	0.0124	0.0065	114	-0.86	027.33	0
027	Magnesium (%)	2220	0.2005	0.0010	0.2112	0.0124	0.0065	114	-0.86	027.44	0
027	Magnesium (%)	0051	0.2027	0.0009	0.2112	0.0124	0.0065	114	-0.68	027.41	0
027	Magnesium (%)	0017	0.2035	0.0010	0.2112	0.0124	0.0065	114	-0.62	027.43	0
027	Magnesium (%)	0049	0.2050	0.0100	0.2112	0.0124	0.0065	114	-0.49	027.41	0
027	Magnesium (%)	0265	0.2050	0.0100	0.2112	0.0124	0.0065	114	-0.49	027.42	0
027	Magnesium (%)	0035	0.2050	0.0100	0.2112	0.0124	0.0065	114	-0.49	027.43	0
027	Magnesium (%)	0505	0.2050	0.0100	0.2112	0.0124	0.0065	114	-0.49	027.43	0
027	Magnesium (%)	2192	0.2050	0.0100	0.2112	0.0124	0.0065	114	-0.49	027.43	0
027	Magnesium (%)	2022	0.2055	0.0010	0.2112	0.0124	0.0065	114	-0.45	027.31	0
027	Magnesium (%)	0083	0.2060	0.0020	0.2112	0.0124	0.0065	114	-0.41	027.43	0
027	Magnesium (%)	2219	0.2060	0.0040	0.2112	0.0124	0.0065	114	-0.41	027.44	0
027	Magnesium (%)	0693	0.2065	0.0050	0.2112	0.0124	0.0065	114	-0.37	027.42	0
027	Magnesium (%)	0572	0.2065	0.0130	0.2112	0.0124	0.0065	114	-0.37	027.53	0
027	Magnesium (%)	2252	0.2070	0.0060	0.2112	0.0124	0.0065	114	-0.33	027.44	0
027	Magnesium (%)	0968	0.2075	0.0030	0.2112	0.0124	0.0065	114	-0.29	027.43	0
027	Magnesium (%)	0142	0.2088	0.0103	0.2112	0.0124	0.0065	114	-0.19	027.31	0
027	Magnesium (%)	0038	0.2090	0.0200	0.2112	0.0124	0.0065	114	-0.17	027.43	0
027	Magnesium (%)	0955	0.2095	0.0050	0.2112	0.0124	0.0065	114	-0.13	027.44	0
027	Magnesium (%)	0263	0.2095	0.0042	0.2112	0.0124	0.0065	114	-0.13	027.42	0
027	Magnesium (%)	0656	0.2100	0.0000	0.2112	0.0124	0.0065	114	-0.09	027.31	0
027	Magnesium (%)	0868	0.2100	0.0000	0.2112	0.0124	0.0065	114	-0.09	027.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
027	Magnesium (%)	0098	0.2100	0.0000	0.2112	0.0124	0.0065	114	-0.09	027.41	0
027	Magnesium (%)	0164	0.2100	0.0020	0.2112	0.0124	0.0065	114	-0.09	027.41	0
027	Magnesium (%)	0520	0.2100	0.0000	0.2112	0.0124	0.0065	114	-0.09	027.41	0
027	Magnesium (%)	0910	0.2100	0.0000	0.2112	0.0124	0.0065	114	-0.09	027.41	0
027	Magnesium (%)	2012	0.2100	0.0000	0.2112	0.0124	0.0065	114	-0.09	027.41	0
027	Magnesium (%)	0675	0.2100	0.0000	0.2112	0.0124	0.0065	114	-0.09	027.43	0
027	Magnesium (%)	0242	0.2100	0.0000	0.2112	0.0124	0.0065	114	-0.09	027.99	0
027	Magnesium (%)	0590	0.2100	0.0000	0.2112	0.0124	0.0065	114	-0.09	027.99	0
027	Magnesium (%)	0003	0.2100	0.0200	0.2112	0.0124	0.0065	114	-0.09	027.41	0
027	Magnesium (%)	0278	0.2100	0.0200	0.2112	0.0124	0.0065	114	-0.09	027.42	0
027	Magnesium (%)	0425	0.2100	0.0200	0.2112	0.0124	0.0065	114	-0.09	027.43	0
027	Magnesium (%)	0874	0.2115	0.0010	0.2112	0.0124	0.0065	114	0.03	027.31	0
027	Magnesium (%)	0139	0.2120	0.0010	0.2112	0.0124	0.0065	114	0.07	027.41	0
027	Magnesium (%)	0918	0.2125	0.0030	0.2112	0.0124	0.0065	114	0.11	027.43	0
027	Magnesium (%)	0074	0.2130	0.0000	0.2112	0.0124	0.0065	114	0.15	027.41	0
027	Magnesium (%)	0045	0.2130	0.0100	0.2112	0.0124	0.0065	114	0.15	027.42	0
027	Magnesium (%)	0065	0.2130	0.0020	0.2112	0.0124	0.0065	114	0.15	027.44	0
027	Magnesium (%)	0208	0.2135	0.0150	0.2112	0.0124	0.0065	114	0.19	027.31	0
027	Magnesium (%)	0186	0.2140	0.0020	0.2112	0.0124	0.0065	114	0.23	027.52	0
027	Magnesium (%)	0560	0.2144	0.0008	0.2112	0.0124	0.0065	114	0.26	027.42	0
027	Magnesium (%)	0619	0.2145	0.0030	0.2112	0.0124	0.0065	114	0.27	027.41	0
027	Magnesium (%)	2114	0.2148	0.0050	0.2112	0.0124	0.0065	114	0.29	027.44	0
027	Magnesium (%)	0536	0.2150	0.0100	0.2112	0.0124	0.0065	114	0.31	027.31	0
027	Magnesium (%)	0035	0.2150	0.0100	0.2112	0.0124	0.0065	114	0.31	027.42	0
027	Magnesium (%)	0190	0.2150	0.0100	0.2112	0.0124	0.0065	114	0.31	027.42	0
027	Magnesium (%)	2207	0.2150	0.0100	0.2112	0.0124	0.0065	114	0.31	027.42	0
027	Magnesium (%)	0033	0.2150	0.0100	0.2112	0.0124	0.0065	114	0.31	027.43	0
027	Magnesium (%)	0098	0.2150	0.0100	0.2112	0.0124	0.0065	114	0.31	027.43	0
027	Magnesium (%)	2124	0.2150	0.0100	0.2112	0.0124	0.0065	114	0.31	027.43	0
027	Magnesium (%)	0613	0.2150	0.0300	0.2112	0.0124	0.0065	114	0.31	027.42	0
027	Magnesium (%)	0171	0.2160	0.0020	0.2112	0.0124	0.0065	114	0.39	027.41	0
027	Magnesium (%)	0208	0.2160	0.0040	0.2112	0.0124	0.0065	114	0.39	027.41	0
027	Magnesium (%)	0726	0.2169	0.0011	0.2112	0.0124	0.0065	114	0.46	027.42	0
027	Magnesium (%)	2115	0.2169	0.0020	0.2112	0.0124	0.0065	114	0.46	027.31	0
027	Magnesium (%)	2156	0.2190	0.0040	0.2112	0.0124	0.0065	114	0.63	027.44	0
027	Magnesium (%)	0695	0.2195	0.0030	0.2112	0.0124	0.0065	114	0.67	027.41	0
027	Magnesium (%)	2196	0.2200	0.0000	0.2112	0.0124	0.0065	114	0.71	027.31	0
027	Magnesium (%)	0169	0.2200	0.0000	0.2112	0.0124	0.0065	114	0.71	027.32	0
027	Magnesium (%)	0019	0.2200	0.0000	0.2112	0.0124	0.0065	114	0.71	027.41	0
027	Magnesium (%)	0354	0.2200	0.0000	0.2112	0.0124	0.0065	114	0.71	027.41	0
027	Magnesium (%)	0405	0.2200	0.0000	0.2112	0.0124	0.0065	114	0.71	027.41	0
027	Magnesium (%)	0878	0.2200	0.0200	0.2112	0.0124	0.0065	114	0.71	027.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
027	Magnesium (%)	0366	0.2200	0.0000	0.2112	0.0124	0.0065	114	0.71	027.42	0
027	Magnesium (%)	0089	0.2200	0.0000	0.2112	0.0124	0.0065	114	0.71	027.43	0
027	Magnesium (%)	0610	0.2200	0.0000	0.2112	0.0124	0.0065	114	0.71	027.43	0
027	Magnesium (%)	0682	0.2200	0.0000	0.2112	0.0124	0.0065	114	0.71	027.43	0
027	Magnesium (%)	0653	0.2210	0.0020	0.2112	0.0124	0.0065	114	0.79	027.41	0
027	Magnesium (%)	0964	0.2215	0.0131	0.2112	0.0124	0.0065	114	0.83	027.43	0
027	Magnesium (%)	0512	0.2216	0.0007	0.2112	0.0124	0.0065	114	0.84	027.43	0
027	Magnesium (%)	2187	0.2224	0.0046	0.2112	0.0124	0.0065	114	0.90	027.43	0
027	Magnesium (%)	0598	0.2225	0.0007	0.2112	0.0124	0.0065	114	0.91	027.41	0
027	Magnesium (%)	0001	0.2234	0.0045	0.2112	0.0124	0.0065	114	0.98	027.31	0
027	Magnesium (%)	0187	0.2245	0.0011	0.2112	0.0124	0.0065	114	1.07	027.42	0
027	Magnesium (%)	0226	0.2250	0.0100	0.2112	0.0124	0.0065	114	1.11	027.41	0
027	Magnesium (%)	0358	0.2250	0.0100	0.2112	0.0124	0.0065	114	1.11	027.41	0
027	Magnesium (%)	0294	0.2250	0.0100	0.2112	0.0124	0.0065	114	1.11	027.42	0
027	Magnesium (%)	2033	0.2250	0.0100	0.2112	0.0124	0.0065	114	1.11	027.43	0
027	Magnesium (%)	0939	0.2250	0.0100	0.2112	0.0124	0.0065	114	1.11	027.53	0
027	Magnesium (%)	0186	0.2270	0.0000	0.2112	0.0124	0.0065	114	1.27	027.42	0
027	Magnesium (%)	0357	0.2285	0.0032	0.2112	0.0124	0.0065	114	1.40	027.42	0
027	Magnesium (%)	2246	0.2300	0.0000	0.2112	0.0124	0.0065	114	1.52	027.31	0
027	Magnesium (%)	0028	0.2350	0.0100	0.2112	0.0124	0.0065	114	1.92	027.43	0
027	Magnesium (%)	0914	0.2360	0.0132	0.2112	0.0124	0.0065	114	2.00	027.43	0
027	Magnesium (%)	2053	0.2400	0.0000	0.2112	0.0124	0.0065	114	2.32	027.42	0
027	Magnesium (%)	0100	0.2450	0.0100	0.2112	0.0124	0.0065	114	2.72	027.99	0
027	Magnesium (%)	0353	0.2500	0.0200	0.2112	0.0124	0.0065	114	3.12	027.43	0
027	Magnesium (%)	0032	0.2550	0.0700	0.2112	0.0124	0.0065	114	3.53	027.43	0
027	Magnesium (%)	0042	0.2615	0.0150	0.2112	0.0124	0.0065	114	4.05	027.42	0
027	Magnesium (%)	2144	426.1	7.990	0.2112	0.0124	0.0065	114	34253.90	027.31	2
028	Manganese (mg / kg (ppm))	2128	10.22	0.0000	55.91	4.152	1.278	98	-11.01	028.32	0
028	Manganese (mg / kg (ppm))	2246	34.92	1.940	55.91	4.152	1.278	98	-5.06	028.34	0
028	Manganese (mg / kg (ppm))	0689	43.50	0.4000	55.91	4.152	1.278	98	-2.99	028.31	0
028	Manganese (mg / kg (ppm))	0505	44.57	0.8400	55.91	4.152	1.278	98	-2.73	028.43	0
028	Manganese (mg / kg (ppm))	2188	45.10	1.400	55.91	4.152	1.278	98	-2.60	028.31	0
028	Manganese (mg / kg (ppm))	2146	45.92	1.830	55.91	4.152	1.278	98	-2.41	028.51	0
028	Manganese (mg / kg (ppm))	0536	46.78	1.450	55.91	4.152	1.278	98	-2.20	028.31	0
028	Manganese (mg / kg (ppm))	0870	47.23	0.4100	55.91	4.152	1.278	98	-2.09	028.42	0
028	Manganese (mg / kg (ppm))	0955	47.40	1.000	55.91	4.152	1.278	98	-2.05	028.44	0
028	Manganese (mg / kg (ppm))	0001	48.71	2.710	55.91	4.152	1.278	98	-1.74	028.31	0
028	Manganese (mg / kg (ppm))	2012	49.01	0.6600	55.91	4.152	1.278	98	-1.66	028.41	0
028	Manganese (mg / kg (ppm))	0695	49.15	0.1170	55.91	4.152	1.278	98	-1.63	028.41	0
028	Manganese (mg / kg (ppm))	0675	50.33	1.940	55.91	4.152	1.278	98	-1.34	028.43	0
028	Manganese (mg / kg (ppm))	0914	50.85	2.300	55.91	4.152	1.278	98	-1.22	028.43	0
028	Manganese (mg / kg (ppm))	2114	51.08	1.023	55.91	4.152	1.278	98	-1.16	028.44	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
028	Manganese (mg / kg (ppm))	0510	51.50	1.000	55.91	4.152	1.278	98	-1.06	028.43	0
028	Manganese (mg / kg (ppm))	0668	52.10	4.200	55.91	4.152	1.278	98	-0.92	028.43	0
028	Manganese (mg / kg (ppm))	2055	52.10	0.2000	55.91	4.152	1.278	98	-0.92	028.43	0
028	Manganese (mg / kg (ppm))	0045	52.40	1.800	55.91	4.152	1.278	98	-0.85	028.42	0
028	Manganese (mg / kg (ppm))	0511	52.50	1.000	55.91	4.152	1.278	98	-0.82	028.41	0
028	Manganese (mg / kg (ppm))	0049	52.71	0.6800	55.91	4.152	1.278	98	-0.77	028.41	0
028	Manganese (mg / kg (ppm))	0693	53.06	2.981	55.91	4.152	1.278	98	-0.69	028.42	0
028	Manganese (mg / kg (ppm))	0169	53.20	0.4000	55.91	4.152	1.278	98	-0.65	028.43	0
028	Manganese (mg / kg (ppm))	0407	53.20	0.4941	55.91	4.152	1.278	98	-0.65	028.41	0
028	Manganese (mg / kg (ppm))	0164	53.50	3.000	55.91	4.152	1.278	98	-0.58	028.41	0
028	Manganese (mg / kg (ppm))	0033	53.65	1.100	55.91	4.152	1.278	98	-0.54	028.43	0
028	Manganese (mg / kg (ppm))	0066	53.83	1.810	55.91	4.152	1.278	98	-0.50	028.44	0
028	Manganese (mg / kg (ppm))	0405	53.98	0.0000	55.91	4.152	1.278	98	-0.46	028.41	0
028	Manganese (mg / kg (ppm))	0610	54.00	4.000	55.91	4.152	1.278	98	-0.46	028.43	0
028	Manganese (mg / kg (ppm))	0171	54.06	0.2400	55.91	4.152	1.278	98	-0.45	028.41	0
028	Manganese (mg / kg (ppm))	0345	54.15	0.1100	55.91	4.152	1.278	98	-0.43	028.43	0
028	Manganese (mg / kg (ppm))	0051	54.40	0.0100	55.91	4.152	1.278	98	-0.36	028.41	0
028	Manganese (mg / kg (ppm))	2115	54.49	0.0600	55.91	4.152	1.278	98	-0.34	028.31	0
028	Manganese (mg / kg (ppm))	0353	54.50	1.000	55.91	4.152	1.278	98	-0.34	028.43	0
028	Manganese (mg / kg (ppm))	0948	54.58	0.5580	55.91	4.152	1.278	98	-0.32	028.33	0
028	Manganese (mg / kg (ppm))	2129	54.64	4.710	55.91	4.152	1.278	98	-0.31	028.99	0
028	Manganese (mg / kg (ppm))	0229	54.69	2.060	55.91	4.152	1.278	98	-0.29	028.41	0
028	Manganese (mg / kg (ppm))	0572	54.70	2.400	55.91	4.152	1.278	98	-0.29	028.53	0
028	Manganese (mg / kg (ppm))	0613	54.85	3.700	55.91	4.152	1.278	98	-0.26	028.42	0
028	Manganese (mg / kg (ppm))	2146	54.89	4.160	55.91	4.152	1.278	98	-0.25	028.31	0
028	Manganese (mg / kg (ppm))	0868	55.05	0.4600	55.91	4.152	1.278	98	-0.21	028.31	0
028	Manganese (mg / kg (ppm))	0265	55.20	0.8000	55.91	4.152	1.278	98	-0.17	028.42	0
028	Manganese (mg / kg (ppm))	0004	55.22	1.260	55.91	4.152	1.278	98	-0.17	028.41	0
028	Manganese (mg / kg (ppm))	0083	55.50	1.000	55.91	4.152	1.278	98	-0.10	028.43	0
028	Manganese (mg / kg (ppm))	0598	55.70	0.0800	55.91	4.152	1.278	98	-0.05	028.41	0
028	Manganese (mg / kg (ppm))	0425	55.93	0.0500	55.91	4.152	1.278	98	0.00	028.43	0
028	Manganese (mg / kg (ppm))	0074	56.00	0.0000	55.91	4.152	1.278	98	0.02	028.41	0
028	Manganese (mg / kg (ppm))	0878	56.00	2.000	55.91	4.152	1.278	98	0.02	028.41	0
028	Manganese (mg / kg (ppm))	0918	56.06	0.2800	55.91	4.152	1.278	98	0.04	028.43	0
028	Manganese (mg / kg (ppm))	0011	56.47	1.278	55.91	4.152	1.278	98	0.13	028.41	0
028	Manganese (mg / kg (ppm))	0910	56.50	1.000	55.91	4.152	1.278	98	0.14	028.41	0
028	Manganese (mg / kg (ppm))	0242	56.50	1.000	55.91	4.152	1.278	98	0.14	028.99	0
028	Manganese (mg / kg (ppm))	0098	56.61	1.240	55.91	4.152	1.278	98	0.17	028.41	0
028	Manganese (mg / kg (ppm))	0619	56.70	0.0000	55.91	4.152	1.278	98	0.19	028.41	0
028	Manganese (mg / kg (ppm))	2192	56.99	0.7900	55.91	4.152	1.278	98	0.26	028.53	0
028	Manganese (mg / kg (ppm))	0278	57.00	0.0000	55.91	4.152	1.278	98	0.26	028.42	0
028	Manganese (mg / kg (ppm))	0263	57.02	0.2160	55.91	4.152	1.278	98	0.27	028.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
028	Manganese (mg / kg (ppm))	0529	57.20	0.2000	55.91	4.152	1.278	98	0.31	028.31	0
028	Manganese (mg / kg (ppm))	2144	57.42	0.8600	55.91	4.152	1.278	98	0.36	028.31	0
028	Manganese (mg / kg (ppm))	0968	57.50	1.000	55.91	4.152	1.278	98	0.38	028.43	0
028	Manganese (mg / kg (ppm))	0186	57.50	1.000	55.91	4.152	1.278	98	0.38	028.52	0
028	Manganese (mg / kg (ppm))	0682	57.54	0.0000	55.91	4.152	1.278	98	0.39	028.43	0
028	Manganese (mg / kg (ppm))	0037	57.75	0.1000	55.91	4.152	1.278	98	0.44	028.42	0
028	Manganese (mg / kg (ppm))	0512	57.92	1.630	55.91	4.152	1.278	98	0.48	028.43	0
028	Manganese (mg / kg (ppm))	0208	57.93	3.030	55.91	4.152	1.278	98	0.49	028.41	0
028	Manganese (mg / kg (ppm))	0047	57.97	2.930	55.91	4.152	1.278	98	0.50	028.52	0
028	Manganese (mg / kg (ppm))	2207	58.00	0.0000	55.91	4.152	1.278	98	0.50	028.42	0
028	Manganese (mg / kg (ppm))	0038	58.05	2.700	55.91	4.152	1.278	98	0.52	028.43	0
028	Manganese (mg / kg (ppm))	2033	58.10	2.000	55.91	4.152	1.278	98	0.53	028.43	0
028	Manganese (mg / kg (ppm))	0656	58.30	2.330	55.91	4.152	1.278	98	0.57	028.31	0
028	Manganese (mg / kg (ppm))	2089	58.43	0.2200	55.91	4.152	1.278	98	0.61	028.43	0
028	Manganese (mg / kg (ppm))	2196	58.50	0.0000	55.91	4.152	1.278	98	0.62	028.31	0
028	Manganese (mg / kg (ppm))	0520	58.50	1.000	55.91	4.152	1.278	98	0.62	028.41	0
028	Manganese (mg / kg (ppm))	0366	58.50	1.000	55.91	4.152	1.278	98	0.62	028.42	0
028	Manganese (mg / kg (ppm))	0939	58.55	0.6400	55.91	4.152	1.278	98	0.64	028.53	0
028	Manganese (mg / kg (ppm))	0019	58.79	2.020	55.91	4.152	1.278	98	0.69	028.41	0
028	Manganese (mg / kg (ppm))	0098	58.79	0.2600	55.91	4.152	1.278	98	0.69	028.43	0
028	Manganese (mg / kg (ppm))	0208	58.90	2.800	55.91	4.152	1.278	98	0.72	028.31	0
028	Manganese (mg / kg (ppm))	0226	58.96	0.7600	55.91	4.152	1.278	98	0.73	028.41	0
028	Manganese (mg / kg (ppm))	0187	59.12	1.170	55.91	4.152	1.278	98	0.77	028.42	0
028	Manganese (mg / kg (ppm))	0560	59.40	1.000	55.91	4.152	1.278	98	0.84	028.42	0
028	Manganese (mg / kg (ppm))	0590	59.40	1.200	55.91	4.152	1.278	98	0.84	028.99	0
028	Manganese (mg / kg (ppm))	0035	59.50	1.000	55.91	4.152	1.278	98	0.86	028.42	0
028	Manganese (mg / kg (ppm))	0202	59.55	0.8970	55.91	4.152	1.278	98	0.88	028.43	0
028	Manganese (mg / kg (ppm))	0964	59.78	1.782	55.91	4.152	1.278	98	0.93	028.43	0
028	Manganese (mg / kg (ppm))	0003	60.00	2.000	55.91	4.152	1.278	98	0.99	028.41	0
028	Manganese (mg / kg (ppm))	0357	60.14	1.076	55.91	4.152	1.278	98	1.02	028.42	0
028	Manganese (mg / kg (ppm))	0354	60.37	0.0500	55.91	4.152	1.278	98	1.07	028.41	0
028	Manganese (mg / kg (ppm))	0596	61.00	0.3200	55.91	4.152	1.278	98	1.23	028.41	0
028	Manganese (mg / kg (ppm))	0100	61.00	2.000	55.91	4.152	1.278	98	1.23	028.99	0
028	Manganese (mg / kg (ppm))	0035	61.50	3.000	55.91	4.152	1.278	98	1.35	028.43	0
028	Manganese (mg / kg (ppm))	2124	61.96	2.060	55.91	4.152	1.278	98	1.46	028.43	0
028	Manganese (mg / kg (ppm))	0726	62.17	0.2900	55.91	4.152	1.278	98	1.51	028.42	0
028	Manganese (mg / kg (ppm))	0190	62.30	0.0100	55.91	4.152	1.278	98	1.54	028.42	0
028	Manganese (mg / kg (ppm))	0358	62.46	3.280	55.91	4.152	1.278	98	1.58	028.41	0
028	Manganese (mg / kg (ppm))	0186	62.50	1.000	55.91	4.152	1.278	98	1.59	028.42	0
028	Manganese (mg / kg (ppm))	0042	63.90	5.400	55.91	4.152	1.278	98	1.92	028.42	0
028	Manganese (mg / kg (ppm))	0294	80.45	0.1000	55.91	4.152	1.278	98	5.91	028.42	0
028	Manganese (mg / kg (ppm))	2053	55.15	7.300	55.91	4.152	1.278	98	-0.18	028.42	1

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
028	Manganese (mg / kg (ppm))	0609	0.1150	0.0100	55.91	4.152	1.278	98	-13.44	028.32	2
031	Phosphorus (%)	0609	0.7150	0.0100	1.082	0.0468	0.0211	148	-7.84	031.01	0
031	Phosphorus (%)	0889	0.7500	0.0000	1.082	0.0468	0.0211	148	-7.09	031.99	0
031	Phosphorus (%)	2031	0.8880	0.0135	1.082	0.0468	0.0211	148	-4.14	031.01	0
031	Phosphorus (%)	0108	0.9500	0.0000	1.082	0.0468	0.0211	148	-2.81	031.01	0
031	Phosphorus (%)	0848	0.9700	0.0000	1.082	0.0468	0.0211	148	-2.39	031.41	0
031	Phosphorus (%)	0154	0.9803	0.0620	1.082	0.0468	0.0211	148	-2.17	031.52	0
031	Phosphorus (%)	0873	0.9850	0.0100	1.082	0.0468	0.0211	148	-2.07	031.44	0
031	Phosphorus (%)	2055	0.9943	0.0438	1.082	0.0468	0.0211	148	-1.87	031.43	0
031	Phosphorus (%)	2181	0.9950	0.0100	1.082	0.0468	0.0211	148	-1.85	031.01	0
031	Phosphorus (%)	0852	0.9950	0.0100	1.082	0.0468	0.0211	148	-1.85	031.99	0
031	Phosphorus (%)	0693	1.004	0.0070	1.082	0.0468	0.0211	148	-1.67	031.42	0
031	Phosphorus (%)	0910	1.005	0.0700	1.082	0.0468	0.0211	148	-1.64	031.41	0
031	Phosphorus (%)	0066	1.005	0.0500	1.082	0.0468	0.0211	148	-1.64	031.44	0
031	Phosphorus (%)	2018	1.010	0.0200	1.082	0.0468	0.0211	148	-1.53	031.01	0
031	Phosphorus (%)	2012	1.015	0.0100	1.082	0.0468	0.0211	148	-1.42	031.41	0
031	Phosphorus (%)	0278	1.015	0.0300	1.082	0.0468	0.0211	148	-1.42	031.42	0
031	Phosphorus (%)	2192	1.030	0.0000	1.082	0.0468	0.0211	148	-1.10	031.01	0
031	Phosphorus (%)	2246	1.030	0.0000	1.082	0.0468	0.0211	148	-1.10	031.01	0
031	Phosphorus (%)	0004	1.030	0.0200	1.082	0.0468	0.0211	148	-1.10	031.41	0
031	Phosphorus (%)	0171	1.030	0.0000	1.082	0.0468	0.0211	148	-1.10	031.41	0
031	Phosphorus (%)	0100	1.030	0.0200	1.082	0.0468	0.0211	148	-1.10	031.99	0
031	Phosphorus (%)	0511	1.035	0.0100	1.082	0.0468	0.0211	148	-1.00	031.01	0
031	Phosphorus (%)	0520	1.035	0.0300	1.082	0.0468	0.0211	148	-1.00	031.41	0
031	Phosphorus (%)	0619	1.035	0.0100	1.082	0.0468	0.0211	148	-1.00	031.41	0
031	Phosphorus (%)	0407	1.039	0.0312	1.082	0.0468	0.0211	148	-0.92	031.41	0
031	Phosphorus (%)	2126	1.040	0.0400	1.082	0.0468	0.0211	148	-0.89	031.01	0
031	Phosphorus (%)	2220	1.040	0.0200	1.082	0.0468	0.0211	148	-0.89	031.01	0
031	Phosphorus (%)	0870	1.043	0.0520	1.082	0.0468	0.0211	148	-0.82	031.42	0
031	Phosphorus (%)	2022	1.045	0.0100	1.082	0.0468	0.0211	148	-0.78	031.01	0
031	Phosphorus (%)	2144	1.045	0.0100	1.082	0.0468	0.0211	148	-0.78	031.01	0
031	Phosphorus (%)	0019	1.045	0.0300	1.082	0.0468	0.0211	148	-0.78	031.41	0
031	Phosphorus (%)	0695	1.045	0.0120	1.082	0.0468	0.0211	148	-0.78	031.41	0
031	Phosphorus (%)	0144	1.050	0.0200	1.082	0.0468	0.0211	148	-0.68	031.41	0
031	Phosphorus (%)	0358	1.050	0.0200	1.082	0.0468	0.0211	148	-0.68	031.41	0
031	Phosphorus (%)	2089	1.050	0.0000	1.082	0.0468	0.0211	148	-0.68	031.43	0
031	Phosphorus (%)	0676	1.050	0.0800	1.082	0.0468	0.0211	148	-0.68	031.99	0
031	Phosphorus (%)	0142	1.053	0.0099	1.082	0.0468	0.0211	148	-0.61	031.01	0
031	Phosphorus (%)	0242	1.055	0.0300	1.082	0.0468	0.0211	148	-0.57	031.99	0
031	Phosphorus (%)	0651	1.057	0.0030	1.082	0.0468	0.0211	148	-0.54	031.01	0
031	Phosphorus (%)	0948	1.057	0.0130	1.082	0.0468	0.0211	148	-0.54	031.01	0
031	Phosphorus (%)	2187	1.058	0.0063	1.082	0.0468	0.0211	148	-0.51	031.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	0590	1.060	0.0000	1.082	0.0468	0.0211	148	-0.46	031.01	0
031	Phosphorus (%)	2252	1.060	0.0200	1.082	0.0468	0.0211	148	-0.46	031.01	0
031	Phosphorus (%)	0169	1.060	0.0000	1.082	0.0468	0.0211	148	-0.46	031.03	0
031	Phosphorus (%)	0164	1.060	0.0200	1.082	0.0468	0.0211	148	-0.46	031.41	0
031	Phosphorus (%)	0298	1.060	0.0200	1.082	0.0468	0.0211	148	-0.46	031.41	0
031	Phosphorus (%)	0045	1.060	0.0600	1.082	0.0468	0.0211	148	-0.46	031.42	0
031	Phosphorus (%)	0065	1.060	0.0200	1.082	0.0468	0.0211	148	-0.46	031.44	0
031	Phosphorus (%)	0139	1.063	0.0050	1.082	0.0468	0.0211	148	-0.41	031.41	0
031	Phosphorus (%)	0726	1.064	0.0080	1.082	0.0468	0.0211	148	-0.38	031.42	0
031	Phosphorus (%)	0885	1.065	0.0010	1.082	0.0468	0.0211	148	-0.36	031.01	0
031	Phosphorus (%)	2146	1.065	0.0500	1.082	0.0468	0.0211	148	-0.35	031.01	0
031	Phosphorus (%)	0010	1.065	0.0100	1.082	0.0468	0.0211	148	-0.35	031.43	0
031	Phosphorus (%)	0345	1.065	0.0100	1.082	0.0468	0.0211	148	-0.35	031.43	0
031	Phosphorus (%)	0229	1.068	0.0340	1.082	0.0468	0.0211	148	-0.29	031.41	0
031	Phosphorus (%)	0560	1.069	0.0156	1.082	0.0468	0.0211	148	-0.28	031.42	0
031	Phosphorus (%)	0674	1.070	0.0000	1.082	0.0468	0.0211	148	-0.25	031.01	0
031	Phosphorus (%)	0689	1.070	0.0200	1.082	0.0468	0.0211	148	-0.25	031.01	0
031	Phosphorus (%)	2188	1.070	0.0200	1.082	0.0468	0.0211	148	-0.25	031.01	0
031	Phosphorus (%)	2196	1.070	0.0000	1.082	0.0468	0.0211	148	-0.25	031.01	0
031	Phosphorus (%)	0033	1.070	0.0800	1.082	0.0468	0.0211	148	-0.25	031.43	0
031	Phosphorus (%)	0955	1.070	0.0200	1.082	0.0468	0.0211	148	-0.25	031.44	0
031	Phosphorus (%)	0914	1.073	0.0089	1.082	0.0468	0.0211	148	-0.18	031.43	0
031	Phosphorus (%)	0656	1.075	0.0100	1.082	0.0468	0.0211	148	-0.14	031.01	0
031	Phosphorus (%)	0897	1.075	0.0100	1.082	0.0468	0.0211	148	-0.14	031.01	0
031	Phosphorus (%)	2009	1.075	0.0010	1.082	0.0468	0.0211	148	-0.13	031.01	0
031	Phosphorus (%)	0098	1.076	0.0320	1.082	0.0468	0.0211	148	-0.12	031.41	0
031	Phosphorus (%)	2114	1.076	0.0145	1.082	0.0468	0.0211	148	-0.12	031.41	0
031	Phosphorus (%)	0849	1.078	0.0050	1.082	0.0468	0.0211	148	-0.09	031.01	0
031	Phosphorus (%)	0687	1.080	0.0400	1.082	0.0468	0.0211	148	-0.03	031.01	0
031	Phosphorus (%)	2058	1.080	0.0000	1.082	0.0468	0.0211	148	-0.03	031.01	0
031	Phosphorus (%)	2193	1.080	0.0000	1.082	0.0468	0.0211	148	-0.03	031.01	0
031	Phosphorus (%)	2219	1.080	0.0000	1.082	0.0468	0.0211	148	-0.03	031.01	0
031	Phosphorus (%)	0003	1.080	0.0000	1.082	0.0468	0.0211	148	-0.03	031.41	0
031	Phosphorus (%)	0354	1.080	0.0000	1.082	0.0468	0.0211	148	-0.03	031.41	0
031	Phosphorus (%)	0425	1.080	0.0200	1.082	0.0468	0.0211	148	-0.03	031.43	0
031	Phosphorus (%)	0510	1.080	0.0000	1.082	0.0468	0.0211	148	-0.03	031.43	0
031	Phosphorus (%)	0939	1.080	0.0200	1.082	0.0468	0.0211	148	-0.03	031.53	0
031	Phosphorus (%)	0037	1.082	0.0230	1.082	0.0468	0.0211	148	0.00	031.42	0
031	Phosphorus (%)	0884	1.085	0.0300	1.082	0.0468	0.0211	148	0.07	031.01	0
031	Phosphorus (%)	0208	1.086	0.0130	1.082	0.0468	0.0211	148	0.08	031.01	0
031	Phosphorus (%)	0263	1.087	0.0060	1.082	0.0468	0.0211	148	0.11	031.42	0
031	Phosphorus (%)	0881	1.087	0.0080	1.082	0.0468	0.0211	148	0.12	031.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	2129	1.090	0.1000	1.082	0.0468	0.0211	148	0.18	031.42	0
031	Phosphorus (%)	0405	1.090	0.0000	1.082	0.0468	0.0211	148	0.18	031.41	0
031	Phosphorus (%)	0186	1.090	0.0200	1.082	0.0468	0.0211	148	0.18	031.52	0
031	Phosphorus (%)	0036	1.090	0.0289	1.082	0.0468	0.0211	148	0.19	031.03	0
031	Phosphorus (%)	0626	1.091	0.0170	1.082	0.0468	0.0211	148	0.19	031.01	0
031	Phosphorus (%)	0202	1.091	0.0560	1.082	0.0468	0.0211	148	0.20	031.43	0
031	Phosphorus (%)	2109	1.093	0.0090	1.082	0.0468	0.0211	148	0.23	031.01	0
031	Phosphorus (%)	0874	1.093	0.0095	1.082	0.0468	0.0211	148	0.25	031.01	0
031	Phosphorus (%)	0918	1.094	0.0290	1.082	0.0468	0.0211	148	0.26	031.43	0
031	Phosphorus (%)	0893	1.095	0.0100	1.082	0.0468	0.0211	148	0.29	031.00	0
031	Phosphorus (%)	0683	1.095	0.0500	1.082	0.0468	0.0211	148	0.29	031.01	0
031	Phosphorus (%)	0868	1.095	0.0300	1.082	0.0468	0.0211	148	0.29	031.01	0
031	Phosphorus (%)	2115	1.095	0.0100	1.082	0.0468	0.0211	148	0.29	031.01	0
031	Phosphorus (%)	0049	1.095	0.0300	1.082	0.0468	0.0211	148	0.29	031.41	0
031	Phosphorus (%)	0878	1.100	0.0200	1.082	0.0468	0.0211	148	0.39	031.01	0
031	Phosphorus (%)	0205	1.100	0.0000	1.082	0.0468	0.0211	148	0.39	031.42	0
031	Phosphorus (%)	0089	1.100	0.0000	1.082	0.0468	0.0211	148	0.39	031.43	0
031	Phosphorus (%)	0682	1.100	0.0000	1.082	0.0468	0.0211	148	0.39	031.43	0
031	Phosphorus (%)	0572	1.100	0.1000	1.082	0.0468	0.0211	148	0.39	031.53	0
031	Phosphorus (%)	2006	1.103	0.0100	1.082	0.0468	0.0211	148	0.46	031.01	0
031	Phosphorus (%)	0083	1.104	0.0010	1.082	0.0468	0.0211	148	0.47	031.43	0
031	Phosphorus (%)	0051	1.104	0.0010	1.082	0.0468	0.0211	148	0.47	031.41	0
031	Phosphorus (%)	2190	1.105	0.0515	1.082	0.0468	0.0211	148	0.50	031.01	0
031	Phosphorus (%)	2053	1.105	0.0100	1.082	0.0468	0.0211	148	0.50	031.02	0
031	Phosphorus (%)	0187	1.105	0.0100	1.082	0.0468	0.0211	148	0.50	031.42	0
031	Phosphorus (%)	0017	1.106	0.0210	1.082	0.0468	0.0211	148	0.51	031.43	0
031	Phosphorus (%)	0001	1.107	0.0030	1.082	0.0468	0.0211	148	0.53	031.03	0
031	Phosphorus (%)	0208	1.107	0.0140	1.082	0.0468	0.0211	148	0.54	031.41	0
031	Phosphorus (%)	0265	1.110	0.0800	1.082	0.0468	0.0211	148	0.61	031.42	0
031	Phosphorus (%)	0366	1.110	0.0400	1.082	0.0468	0.0211	148	0.61	031.42	0
031	Phosphorus (%)	0505	1.110	0.0600	1.082	0.0468	0.0211	148	0.61	031.43	0
031	Phosphorus (%)	0035	1.110	0.0200	1.082	0.0468	0.0211	148	0.61	031.42	0
031	Phosphorus (%)	2198	1.115	0.0100	1.082	0.0468	0.0211	148	0.71	031.01	0
031	Phosphorus (%)	0294	1.115	0.0100	1.082	0.0468	0.0211	148	0.71	031.42	0
031	Phosphorus (%)	0968	1.116	0.0080	1.082	0.0468	0.0211	148	0.74	031.43	0
031	Phosphorus (%)	0011	1.118	0.0330	1.082	0.0468	0.0211	148	0.77	031.41	0
031	Phosphorus (%)	0675	1.120	0.0000	1.082	0.0468	0.0211	148	0.82	031.43	0
031	Phosphorus (%)	0036	1.124	0.0028	1.082	0.0468	0.0211	148	0.91	031.43	0
031	Phosphorus (%)	0074	1.125	0.0100	1.082	0.0468	0.0211	148	0.93	031.41	0
031	Phosphorus (%)	2156	1.126	0.0060	1.082	0.0468	0.0211	148	0.95	031.01	0
031	Phosphorus (%)	0653	1.126	0.0080	1.082	0.0468	0.0211	148	0.95	031.41	0
031	Phosphorus (%)	0529	1.130	0.0200	1.082	0.0468	0.0211	148	1.04	031.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	0964	1.135	0.0488	1.082	0.0468	0.0211	148	1.13	031.43	0
031	Phosphorus (%)	0598	1.135	0.0140	1.082	0.0468	0.0211	148	1.14	031.41	0
031	Phosphorus (%)	0357	1.138	0.0084	1.082	0.0468	0.0211	148	1.20	031.42	0
031	Phosphorus (%)	0186	1.140	0.0200	1.082	0.0468	0.0211	148	1.25	031.42	0
031	Phosphorus (%)	2207	1.140	0.0200	1.082	0.0468	0.0211	148	1.25	031.42	0
031	Phosphorus (%)	0610	1.140	0.0200	1.082	0.0468	0.0211	148	1.25	031.43	0
031	Phosphorus (%)	2058	1.142	0.0430	1.082	0.0468	0.0211	148	1.28	031.43	0
031	Phosphorus (%)	0512	1.143	0.0360	1.082	0.0468	0.0211	148	1.31	031.43	0
031	Phosphorus (%)	2033	1.145	0.0300	1.082	0.0468	0.0211	148	1.36	031.43	0
031	Phosphorus (%)	0226	1.150	0.0200	1.082	0.0468	0.0211	148	1.46	031.41	0
031	Phosphorus (%)	2259	1.150	0.0400	1.082	0.0468	0.0211	148	1.46	031.42	0
031	Phosphorus (%)	0860	1.151	0.0000	1.082	0.0468	0.0211	148	1.48	031.43	0
031	Phosphorus (%)	0098	1.151	0.0020	1.082	0.0468	0.0211	148	1.48	031.43	0
031	Phosphorus (%)	2259	1.155	0.0100	1.082	0.0468	0.0211	148	1.57	031.41	0
031	Phosphorus (%)	0190	1.165	0.0100	1.082	0.0468	0.0211	148	1.78	031.42	0
031	Phosphorus (%)	0353	1.165	0.0100	1.082	0.0468	0.0211	148	1.78	031.43	0
031	Phosphorus (%)	0042	1.180	0.0600	1.082	0.0468	0.0211	148	2.10	031.42	0
031	Phosphorus (%)	2124	1.185	0.0100	1.082	0.0468	0.0211	148	2.21	031.43	0
031	Phosphorus (%)	2222	1.190	0.0400	1.082	0.0468	0.0211	148	2.32	031.01	0
031	Phosphorus (%)	0028	1.200	0.0800	1.082	0.0468	0.0211	148	2.53	031.43	0
031	Phosphorus (%)	0035	1.200	0.0400	1.082	0.0468	0.0211	148	2.53	031.43	0
031	Phosphorus (%)	0038	1.210	0.0500	1.082	0.0468	0.0211	148	2.75	031.43	0
031	Phosphorus (%)	0032	1.245	0.0900	1.082	0.0468	0.0211	148	3.50	031.43	0
031	Phosphorus (%)	0668	1.050	0.1400	1.082	0.0468	0.0211	148	-0.68	031.43	1
031	Phosphorus (%)	0613	1.105	0.1500	1.082	0.0468	0.0211	148	0.50	031.42	1
031	Phosphorus (%)	2128	1.268	0.1440	1.082	0.0468	0.0211	148	3.99	031.01	1
032	Potassium (%)	0873	0.5600	0.0400	0.8188	0.0496	0.0187	112	-5.21	032.44	0
032	Potassium (%)	0889	0.6000	0.0000	0.8188	0.0496	0.0187	112	-4.41	032.99	0
032	Potassium (%)	0536	0.6050	0.0100	0.8188	0.0496	0.0187	112	-4.31	032.31	0
032	Potassium (%)	0609	0.6550	0.0500	0.8188	0.0496	0.0187	112	-3.30	032.32	0
032	Potassium (%)	0154	0.6772	0.0433	0.8188	0.0496	0.0187	112	-2.85	032.52	0
032	Potassium (%)	2187	0.7456	0.0242	0.8188	0.0496	0.0187	112	-1.47	032.43	0
032	Potassium (%)	2246	0.7490	0.0027	0.8188	0.0496	0.0187	112	-1.41	032.32	0
032	Potassium (%)	0511	0.7500	0.0200	0.8188	0.0496	0.0187	112	-1.39	032.41	0
032	Potassium (%)	2022	0.7505	0.0170	0.8188	0.0496	0.0187	112	-1.38	032.31	0
032	Potassium (%)	0619	0.7544	0.0007	0.8188	0.0496	0.0187	112	-1.30	032.41	0
032	Potassium (%)	0425	0.7550	0.0100	0.8188	0.0496	0.0187	112	-1.28	032.43	0
032	Potassium (%)	0695	0.7570	0.0180	0.8188	0.0496	0.0187	112	-1.24	032.41	0
032	Potassium (%)	0229	0.7590	0.0360	0.8188	0.0496	0.0187	112	-1.20	032.41	0
032	Potassium (%)	0004	0.7600	0.0200	0.8188	0.0496	0.0187	112	-1.18	032.41	0
032	Potassium (%)	0019	0.7600	0.0000	0.8188	0.0496	0.0187	112	-1.18	032.41	0
032	Potassium (%)	0049	0.7600	0.0200	0.8188	0.0496	0.0187	112	-1.18	032.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
032	Potassium (%)	0693	0.7630	0.0440	0.8188	0.0496	0.0187	112	-1.12	032.42	0
032	Potassium (%)	2114	0.7684	0.0035	0.8188	0.0496	0.0187	112	-1.02	032.44	0
032	Potassium (%)	2129	0.7698	0.0446	0.8188	0.0496	0.0187	112	-0.99	032.42	0
032	Potassium (%)	2220	0.7700	0.0000	0.8188	0.0496	0.0187	112	-0.98	032.44	0
032	Potassium (%)	2055	0.7719	0.0291	0.8188	0.0496	0.0187	112	-0.95	032.43	0
032	Potassium (%)	0045	0.7745	0.0050	0.8188	0.0496	0.0187	112	-0.89	032.42	0
032	Potassium (%)	0868	0.7750	0.0300	0.8188	0.0496	0.0187	112	-0.88	032.31	0
032	Potassium (%)	0003	0.7750	0.0700	0.8188	0.0496	0.0187	112	-0.88	032.41	0
032	Potassium (%)	0354	0.7800	0.0200	0.8188	0.0496	0.0187	112	-0.78	032.41	0
032	Potassium (%)	0278	0.7800	0.0000	0.8188	0.0496	0.0187	112	-0.78	032.42	0
032	Potassium (%)	0948	0.7815	0.0130	0.8188	0.0496	0.0187	112	-0.75	032.31	0
032	Potassium (%)	0914	0.7821	0.0221	0.8188	0.0496	0.0187	112	-0.74	032.43	0
032	Potassium (%)	0874	0.7865	0.0070	0.8188	0.0496	0.0187	112	-0.65	032.31	0
032	Potassium (%)	0011	0.7871	0.0334	0.8188	0.0496	0.0187	112	-0.64	032.41	0
032	Potassium (%)	0870	0.7894	0.0476	0.8188	0.0496	0.0187	112	-0.59	032.42	0
032	Potassium (%)	0144	0.7900	0.0400	0.8188	0.0496	0.0187	112	-0.58	032.41	0
032	Potassium (%)	0682	0.7900	0.0000	0.8188	0.0496	0.0187	112	-0.58	032.43	0
032	Potassium (%)	2089	0.7900	0.0200	0.8188	0.0496	0.0187	112	-0.58	032.43	0
032	Potassium (%)	0668	0.7915	0.0550	0.8188	0.0496	0.0187	112	-0.55	032.43	0
032	Potassium (%)	2058	0.7930	0.0040	0.8188	0.0496	0.0187	112	-0.52	032.43	0
032	Potassium (%)	0675	0.7950	0.0300	0.8188	0.0496	0.0187	112	-0.48	032.43	0
032	Potassium (%)	0139	0.7975	0.0050	0.8188	0.0496	0.0187	112	-0.43	032.41	0
032	Potassium (%)	2219	0.8000	0.0200	0.8188	0.0496	0.0187	112	-0.38	032.44	0
032	Potassium (%)	2252	0.8000	0.0200	0.8188	0.0496	0.0187	112	-0.38	032.44	0
032	Potassium (%)	0955	0.8020	0.0380	0.8188	0.0496	0.0187	112	-0.34	032.44	0
032	Potassium (%)	0202	0.8030	0.0140	0.8188	0.0496	0.0187	112	-0.32	032.43	0
032	Potassium (%)	2012	0.8050	0.0100	0.8188	0.0496	0.0187	112	-0.28	032.41	0
032	Potassium (%)	0208	0.8065	0.0090	0.8188	0.0496	0.0187	112	-0.25	032.41	0
032	Potassium (%)	0345	0.8065	0.0070	0.8188	0.0496	0.0187	112	-0.25	032.43	0
032	Potassium (%)	2115	0.8068	0.0157	0.8188	0.0496	0.0187	112	-0.24	032.31	0
032	Potassium (%)	0083	0.8105	0.0070	0.8188	0.0496	0.0187	112	-0.17	032.43	0
032	Potassium (%)	0358	0.8150	0.0300	0.8188	0.0496	0.0187	112	-0.08	032.41	0
032	Potassium (%)	0510	0.8150	0.0100	0.8188	0.0496	0.0187	112	-0.08	032.43	0
032	Potassium (%)	0208	0.8180	0.0060	0.8188	0.0496	0.0187	112	-0.02	032.31	0
032	Potassium (%)	0164	0.8180	0.0140	0.8188	0.0496	0.0187	112	-0.02	032.41	0
032	Potassium (%)	0037	0.8180	0.0400	0.8188	0.0496	0.0187	112	-0.02	032.42	0
032	Potassium (%)	0918	0.8200	0.0140	0.8188	0.0496	0.0187	112	0.02	032.43	0
032	Potassium (%)	0098	0.8200	0.0200	0.8188	0.0496	0.0187	112	0.02	032.41	0
032	Potassium (%)	0033	0.8200	0.0400	0.8188	0.0496	0.0187	112	0.02	032.43	0
032	Potassium (%)	0171	0.8210	0.0020	0.8188	0.0496	0.0187	112	0.04	032.41	0
032	Potassium (%)	0726	0.8225	0.0009	0.8188	0.0496	0.0187	112	0.07	032.42	0
032	Potassium (%)	0512	0.8234	0.0114	0.8188	0.0496	0.0187	112	0.09	032.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
032	Potassium (%)	0520	0.8250	0.0100	0.8188	0.0496	0.0187	112	0.13	032.41	0
032	Potassium (%)	0098	0.8250	0.0300	0.8188	0.0496	0.0187	112	0.13	032.43	0
032	Potassium (%)	0505	0.8250	0.0100	0.8188	0.0496	0.0187	112	0.13	032.43	0
032	Potassium (%)	0100	0.8250	0.0100	0.8188	0.0496	0.0187	112	0.13	032.99	0
032	Potassium (%)	0187	0.8257	0.0041	0.8188	0.0496	0.0187	112	0.14	032.42	0
032	Potassium (%)	0205	0.8280	0.0020	0.8188	0.0496	0.0187	112	0.19	032.42	0
032	Potassium (%)	0683	0.8300	0.0000	0.8188	0.0496	0.0187	112	0.23	032.31	0
032	Potassium (%)	0405	0.8300	0.0000	0.8188	0.0496	0.0187	112	0.23	032.41	0
032	Potassium (%)	0878	0.8300	0.0200	0.8188	0.0496	0.0187	112	0.23	032.41	0
032	Potassium (%)	0065	0.8300	0.0200	0.8188	0.0496	0.0187	112	0.23	032.44	0
032	Potassium (%)	0968	0.8320	0.0040	0.8188	0.0496	0.0187	112	0.27	032.43	0
032	Potassium (%)	0051	0.8332	0.0003	0.8188	0.0496	0.0187	112	0.29	032.41	0
032	Potassium (%)	0910	0.8350	0.0100	0.8188	0.0496	0.0187	112	0.33	032.41	0
032	Potassium (%)	0613	0.8350	0.0700	0.8188	0.0496	0.0187	112	0.33	032.42	0
032	Potassium (%)	2207	0.8350	0.0100	0.8188	0.0496	0.0187	112	0.33	032.42	0
032	Potassium (%)	0035	0.8350	0.0500	0.8188	0.0496	0.0187	112	0.33	032.43	0
032	Potassium (%)	0001	0.8350	0.0400	0.8188	0.0496	0.0187	112	0.33	032.99	0
032	Potassium (%)	0590	0.8350	0.0700	0.8188	0.0496	0.0187	112	0.33	032.99	0
032	Potassium (%)	0142	0.8368	0.0002	0.8188	0.0496	0.0187	112	0.36	032.31	0
032	Potassium (%)	0964	0.8394	0.0526	0.8188	0.0496	0.0187	112	0.42	032.43	0
032	Potassium (%)	0035	0.8400	0.0200	0.8188	0.0496	0.0187	112	0.43	032.42	0
032	Potassium (%)	0186	0.8405	0.0010	0.8188	0.0496	0.0187	112	0.44	032.52	0
032	Potassium (%)	0074	0.8450	0.0500	0.8188	0.0496	0.0187	112	0.53	032.41	0
032	Potassium (%)	0529	0.8500	0.0000	0.8188	0.0496	0.0187	112	0.63	032.31	0
032	Potassium (%)	0038	0.8540	0.0040	0.8188	0.0496	0.0187	112	0.71	032.43	0
032	Potassium (%)	0560	0.8541	0.0019	0.8188	0.0496	0.0187	112	0.71	032.42	0
032	Potassium (%)	0042	0.8545	0.0230	0.8188	0.0496	0.0187	112	0.72	032.42	0
032	Potassium (%)	0407	0.8547	0.0008	0.8188	0.0496	0.0187	112	0.72	032.41	0
032	Potassium (%)	2156	0.8550	0.0120	0.8188	0.0496	0.0187	112	0.73	032.44	0
032	Potassium (%)	0017	0.8560	0.0200	0.8188	0.0496	0.0187	112	0.75	032.43	0
032	Potassium (%)	2053	0.8600	0.0400	0.8188	0.0496	0.0187	112	0.83	032.31	0
032	Potassium (%)	0169	0.8600	0.0200	0.8188	0.0496	0.0187	112	0.83	032.32	0
032	Potassium (%)	0265	0.8600	0.0400	0.8188	0.0496	0.0187	112	0.83	032.42	0
032	Potassium (%)	0089	0.8600	0.0000	0.8188	0.0496	0.0187	112	0.83	032.43	0
032	Potassium (%)	0226	0.8650	0.0100	0.8188	0.0496	0.0187	112	0.93	032.41	0
032	Potassium (%)	0186	0.8675	0.0190	0.8188	0.0496	0.0187	112	0.98	032.42	0
032	Potassium (%)	0653	0.8685	0.0150	0.8188	0.0496	0.0187	112	1.00	032.41	0
032	Potassium (%)	0066	0.8705	0.0110	0.8188	0.0496	0.0187	112	1.04	032.41	0
032	Potassium (%)	0263	0.8745	0.0037	0.8188	0.0496	0.0187	112	1.12	032.42	0
032	Potassium (%)	2216	0.8750	0.0100	0.8188	0.0496	0.0187	112	1.13	032.32	0
032	Potassium (%)	0610	0.8750	0.0100	0.8188	0.0496	0.0187	112	1.13	032.43	0
032	Potassium (%)	2033	0.8750	0.0300	0.8188	0.0496	0.0187	112	1.13	032.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
032	Potassium (%)	0598	0.8845	0.0042	0.8188	0.0496	0.0187	112	1.32	032.41	0
032	Potassium (%)	2124	0.8850	0.0100	0.8188	0.0496	0.0187	112	1.33	032.43	0
032	Potassium (%)	0939	0.8850	0.0300	0.8188	0.0496	0.0187	112	1.33	032.53	0
032	Potassium (%)	0294	0.8900	0.0200	0.8188	0.0496	0.0187	112	1.43	032.42	0
032	Potassium (%)	0572	0.8945	0.0130	0.8188	0.0496	0.0187	112	1.53	032.53	0
032	Potassium (%)	0242	0.8950	0.0100	0.8188	0.0496	0.0187	112	1.54	032.99	0
032	Potassium (%)	0190	0.9100	0.0200	0.8188	0.0496	0.0187	112	1.84	032.42	0
032	Potassium (%)	0353	0.9100	0.0200	0.8188	0.0496	0.0187	112	1.84	032.43	0
032	Potassium (%)	0357	0.9179	0.0029	0.8188	0.0496	0.0187	112	2.00	032.42	0
032	Potassium (%)	0028	0.9250	0.0100	0.8188	0.0496	0.0187	112	2.14	032.43	0
032	Potassium (%)	0366	0.9500	0.0400	0.8188	0.0496	0.0187	112	2.64	032.42	0
032	Potassium (%)	2109	0.9930	0.0000	0.8188	0.0496	0.0187	112	3.51	032.08	0
032	Potassium (%)	0032	0.7550	0.0900	0.8188	0.0496	0.0187	112	-1.28	032.43	1
033	Salt as chloride (%)	0366	0.4400	0.0800	0.7753	0.0771	0.0214	60	-4.35	033.00	0
033	Salt as chloride (%)	0656	0.6050	0.0900	0.7753	0.0771	0.0214	60	-2.21	033.03	0
033	Salt as chloride (%)	0881	0.6060	0.0120	0.7753	0.0771	0.0214	60	-2.20	033.00	0
033	Salt as chloride (%)	0893	0.6400	0.0200	0.7753	0.0771	0.0214	60	-1.75	033.00	0
033	Salt as chloride (%)	0171	0.6600	0.0000	0.7753	0.0771	0.0214	60	-1.49	033.00	0
033	Salt as chloride (%)	0675	0.6600	0.0400	0.7753	0.0771	0.0214	60	-1.49	033.00	0
033	Salt as chloride (%)	0171	0.6600	0.0000	0.7753	0.0771	0.0214	60	-1.49	033.05	0
033	Salt as chloride (%)	0897	0.6750	0.0100	0.7753	0.0771	0.0214	60	-1.30	033.00	0
033	Salt as chloride (%)	0885	0.6867	0.0079	0.7753	0.0771	0.0214	60	-1.15	033.00	0
033	Salt as chloride (%)	2196	0.6900	0.0000	0.7753	0.0771	0.0214	60	-1.11	033.00	0
033	Salt as chloride (%)	2073	0.6900	0.0600	0.7753	0.0771	0.0214	60	-1.11	033.99	0
033	Salt as chloride (%)	0504	0.7150	0.0100	0.7753	0.0771	0.0214	60	-0.78	033.00	0
033	Salt as chloride (%)	2146	0.7250	0.0300	0.7753	0.0771	0.0214	60	-0.65	033.99	0
033	Salt as chloride (%)	0045	0.7275	0.0350	0.7753	0.0771	0.0214	60	-0.62	033.00	0
033	Salt as chloride (%)	0298	0.7300	0.0200	0.7753	0.0771	0.0214	60	-0.59	033.00	0
033	Salt as chloride (%)	0619	0.7325	0.0090	0.7753	0.0771	0.0214	60	-0.55	033.01	0
033	Salt as chloride (%)	0884	0.7450	0.0100	0.7753	0.0771	0.0214	60	-0.39	033.00	0
033	Salt as chloride (%)	0910	0.7450	0.0100	0.7753	0.0771	0.0214	60	-0.39	033.99	0
033	Salt as chloride (%)	0878	0.7500	0.0200	0.7753	0.0771	0.0214	60	-0.33	033.00	0
033	Salt as chloride (%)	2022	0.7500	0.0040	0.7753	0.0771	0.0214	60	-0.33	033.00	0
033	Salt as chloride (%)	2144	0.7500	0.0000	0.7753	0.0771	0.0214	60	-0.33	033.99	0
033	Salt as chloride (%)	0083	0.7600	0.0000	0.7753	0.0771	0.0214	60	-0.20	033.01	0
033	Salt as chloride (%)	0638	0.7600	0.0000	0.7753	0.0771	0.0214	60	-0.20	033.01	0
033	Salt as chloride (%)	0650	0.7600	0.0000	0.7753	0.0771	0.0214	60	-0.20	033.01	0
033	Salt as chloride (%)	0689	0.7650	0.0100	0.7753	0.0771	0.0214	60	-0.13	033.05	0
033	Salt as chloride (%)	0353	0.7700	0.0000	0.7753	0.0771	0.0214	60	-0.07	033.00	0
033	Salt as chloride (%)	0682	0.7700	0.0000	0.7753	0.0771	0.0214	60	-0.07	033.99	0
033	Salt as chloride (%)	0098	0.7750	0.0100	0.7753	0.0771	0.0214	60	0.00	033.01	0
033	Salt as chloride (%)	0354	0.7750	0.0100	0.7753	0.0771	0.0214	60	0.00	033.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
033	Salt as chloride (%)	0510	0.7750	0.0000	0.7753	0.0771	0.0214	60	0.00	033.01	0
033	Salt as chloride (%)	0425	0.7800	0.0000	0.7753	0.0771	0.0214	60	0.06	033.01	0
033	Salt as chloride (%)	0948	0.7850	0.0100	0.7753	0.0771	0.0214	60	0.13	033.01	0
033	Salt as chloride (%)	2146	0.7850	0.0100	0.7753	0.0771	0.0214	60	0.13	033.01	0
033	Salt as chloride (%)	0407	0.7860	0.0020	0.7753	0.0771	0.0214	60	0.14	033.01	0
033	Salt as chloride (%)	0100	0.7900	0.0000	0.7753	0.0771	0.0214	60	0.19	033.01	0
033	Salt as chloride (%)	0683	0.7900	0.0200	0.7753	0.0771	0.0214	60	0.19	033.01	0
033	Salt as chloride (%)	2109	0.7925	0.0230	0.7753	0.0771	0.0214	60	0.22	033.01	0
033	Salt as chloride (%)	0309	0.7950	0.0233	0.7753	0.0771	0.0214	60	0.26	033.00	0
033	Salt as chloride (%)	0164	0.8000	0.0000	0.7753	0.0771	0.0214	60	0.32	033.01	0
033	Salt as chloride (%)	0205	0.8020	0.0000	0.7753	0.0771	0.0214	60	0.35	033.01	0
033	Salt as chloride (%)	0644	0.8030	0.0000	0.7753	0.0771	0.0214	60	0.36	033.01	0
033	Salt as chloride (%)	0868	0.8050	0.0100	0.7753	0.0771	0.0214	60	0.39	033.05	0
033	Salt as chloride (%)	2193	0.8100	0.0600	0.7753	0.0771	0.0214	60	0.45	033.00	0
033	Salt as chloride (%)	0229	0.8100	0.0040	0.7753	0.0771	0.0214	60	0.45	033.01	0
033	Salt as chloride (%)	0186	0.8230	0.0160	0.7753	0.0771	0.0214	60	0.62	033.01	0
033	Salt as chloride (%)	0874	0.8250	0.0000	0.7753	0.0771	0.0214	60	0.65	033.01	0
033	Salt as chloride (%)	0693	0.8265	0.1490	0.7753	0.0771	0.0214	60	0.66	033.00	0
033	Salt as chloride (%)	0265	0.8350	0.0100	0.7753	0.0771	0.0214	60	0.77	033.03	0
033	Salt as chloride (%)	2181	0.8350	0.0500	0.7753	0.0771	0.0214	60	0.77	033.99	0
033	Salt as chloride (%)	2043	0.8400	0.0200	0.7753	0.0771	0.0214	60	0.84	033.05	0
033	Salt as chloride (%)	0242	0.8450	0.1100	0.7753	0.0771	0.0214	60	0.90	033.01	0
033	Salt as chloride (%)	0278	0.8450	0.0300	0.7753	0.0771	0.0214	60	0.90	033.01	0
033	Salt as chloride (%)	2190	0.8552	0.0257	0.7753	0.0771	0.0214	60	1.04	033.00	0
033	Salt as chloride (%)	0017	0.8790	0.0060	0.7753	0.0771	0.0214	60	1.35	033.99	0
033	Salt as chloride (%)	0539	0.8800	0.0600	0.7753	0.0771	0.0214	60	1.36	033.00	0
033	Salt as chloride (%)	0226	0.8900	0.0400	0.7753	0.0771	0.0214	60	1.49	033.01	0
033	Salt as chloride (%)	0358	0.9000	0.0000	0.7753	0.0771	0.0214	60	1.62	033.99	0
033	Salt as chloride (%)	2076	0.9100	0.0000	0.7753	0.0771	0.0214	60	1.75	033.00	0
033	Salt as chloride (%)	0590	0.9200	0.0000	0.7753	0.0771	0.0214	60	1.88	033.01	0
033	Salt as chloride (%)	0144	1.075	0.1100	0.7753	0.0771	0.0214	60	3.89	033.03	0
033	Salt as chloride (%)	0853	1.295	0.8100	0.7753	0.0771	0.0214	60	6.74	033.99	1
034	Selenium (mg / kg (ppm))	0164	0.3365	0.0030	0.4861	0.1268	0.0311	23	-1.18	034.53	0
034	Selenium (mg / kg (ppm))	0964	0.3384	0.1191	0.4861	0.1268	0.0311	23	-1.16	034.43	0
034	Selenium (mg / kg (ppm))	0939	0.3600	0.0200	0.4861	0.1268	0.0311	23	-0.99	034.53	0
034	Selenium (mg / kg (ppm))	2055	0.3800	0.0200	0.4861	0.1268	0.0311	23	-0.84	034.53	0
034	Selenium (mg / kg (ppm))	0045	0.3945	0.0010	0.4861	0.1268	0.0311	23	-0.72	034.04	0
034	Selenium (mg / kg (ppm))	0171	0.3965	0.0310	0.4861	0.1268	0.0311	23	-0.71	034.04	0
034	Selenium (mg / kg (ppm))	0186	0.3990	0.1040	0.4861	0.1268	0.0311	23	-0.69	034.52	0
034	Selenium (mg / kg (ppm))	0047	0.4090	0.0020	0.4861	0.1268	0.0311	23	-0.61	034.52	0
034	Selenium (mg / kg (ppm))	0619	0.4375	0.0150	0.4861	0.1268	0.0311	23	-0.38	034.41	0
034	Selenium (mg / kg (ppm))	0003	0.4460	0.0080	0.4861	0.1268	0.0311	23	-0.32	034.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
034	Selenium (mg / kg (ppm))	0560	0.4515	0.0110	0.4861	0.1268	0.0311	23	-0.27	034.53	0
034	Selenium (mg / kg (ppm))	0038	0.4535	0.0010	0.4861	0.1268	0.0311	23	-0.26	034.01	0
034	Selenium (mg / kg (ppm))	0169	0.4600	0.0000	0.4861	0.1268	0.0311	23	-0.21	034.04	0
034	Selenium (mg / kg (ppm))	2033	0.4950	0.0300	0.4861	0.1268	0.0311	23	0.07	034.43	0
034	Selenium (mg / kg (ppm))	0910	0.5050	0.0100	0.4861	0.1268	0.0311	23	0.15	034.52	0
034	Selenium (mg / kg (ppm))	0208	0.5090	0.0720	0.4861	0.1268	0.0311	23	0.18	034.52	0
034	Selenium (mg / kg (ppm))	2207	0.5400	0.0600	0.4861	0.1268	0.0311	23	0.43	034.52	0
034	Selenium (mg / kg (ppm))	2012	0.5745	0.0190	0.4861	0.1268	0.0311	23	0.70	034.53	0
034	Selenium (mg / kg (ppm))	0918	0.5888	0.0394	0.4861	0.1268	0.0311	23	0.81	034.53	0
034	Selenium (mg / kg (ppm))	0968	0.7970	0.0000	0.4861	0.1268	0.0311	23	2.45	034.43	0
034	Selenium (mg / kg (ppm))	0407	1.440	0.1100	0.4861	0.1268	0.0311	23	7.52	034.41	0
034	Selenium (mg / kg (ppm))	0032	1.610	0.0200	0.4861	0.1268	0.0311	23	8.86	034.53	0
034	Selenium (mg / kg (ppm))	0190	1.610	0.0200	0.4861	0.1268	0.0311	23	8.86	034.99	0
034	Selenium (mg / kg (ppm))	2124	< 1		0.4861	0.1268	0.0311	23		034.43	5
035	Sodium (%)	0873	0.2150	0.0100	0.3445	0.0186	0.0081	117	-6.96	035.41	0
035	Sodium (%)	0609	0.2500	0.0000	0.3445	0.0186	0.0081	117	-5.08	035.32	0
035	Sodium (%)	2006	0.2940	0.0020	0.3445	0.0186	0.0081	117	-2.71	035.01	0
035	Sodium (%)	0536	0.2950	0.0100	0.3445	0.0186	0.0081	117	-2.66	035.31	0
035	Sodium (%)	0037	0.3000	0.0020	0.3445	0.0186	0.0081	117	-2.39	035.05	0
035	Sodium (%)	2192	0.3050	0.0100	0.3445	0.0186	0.0081	117	-2.12	035.43	0
035	Sodium (%)	2196	0.3100	0.0000	0.3445	0.0186	0.0081	117	-1.85	035.31	0
035	Sodium (%)	0066	0.3150	0.0100	0.3445	0.0186	0.0081	117	-1.58	035.41	0
035	Sodium (%)	0505	0.3150	0.0100	0.3445	0.0186	0.0081	117	-1.58	035.43	0
035	Sodium (%)	0693	0.3155	0.0090	0.3445	0.0186	0.0081	117	-1.56	035.42	0
035	Sodium (%)	0298	0.3200	0.0200	0.3445	0.0186	0.0081	117	-1.32	035.41	0
035	Sodium (%)	0154	0.3207	0.0142	0.3445	0.0186	0.0081	117	-1.28	035.52	0
035	Sodium (%)	0366	0.3220	0.0180	0.3445	0.0186	0.0081	117	-1.21	035.42	0
035	Sodium (%)	0868	0.3245	0.0010	0.3445	0.0186	0.0081	117	-1.07	035.01	0
035	Sodium (%)	2220	0.3250	0.0100	0.3445	0.0186	0.0081	117	-1.05	035.41	0
035	Sodium (%)	0572	0.3275	0.0130	0.3445	0.0186	0.0081	117	-0.91	035.53	0
035	Sodium (%)	0668	0.3280	0.0260	0.3445	0.0186	0.0081	117	-0.89	035.43	0
035	Sodium (%)	0914	0.3280	0.0190	0.3445	0.0186	0.0081	117	-0.89	035.43	0
035	Sodium (%)	0695	0.3285	0.0070	0.3445	0.0186	0.0081	117	-0.86	035.41	0
035	Sodium (%)	0345	0.3285	0.0050	0.3445	0.0186	0.0081	117	-0.86	035.43	0
035	Sodium (%)	0870	0.3287	0.0067	0.3445	0.0186	0.0081	117	-0.85	035.42	0
035	Sodium (%)	0171	0.3300	0.0000	0.3445	0.0186	0.0081	117	-0.78	035.41	0
035	Sodium (%)	2012	0.3300	0.0000	0.3445	0.0186	0.0081	117	-0.78	035.41	0
035	Sodium (%)	0425	0.3300	0.0000	0.3445	0.0186	0.0081	117	-0.78	035.43	0
035	Sodium (%)	0675	0.3300	0.0200	0.3445	0.0186	0.0081	117	-0.78	035.43	0
035	Sodium (%)	2089	0.3300	0.0000	0.3445	0.0186	0.0081	117	-0.78	035.43	0
035	Sodium (%)	2129	0.3308	0.0137	0.3445	0.0186	0.0081	117	-0.74	035.99	0
035	Sodium (%)	2114	0.3323	0.0086	0.3445	0.0186	0.0081	117	-0.65	035.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
035	Sodium (%)	2058	0.3345	0.0070	0.3445	0.0186	0.0081	117	-0.54	035.43	0
035	Sodium (%)	0683	0.3350	0.0300	0.3445	0.0186	0.0081	117	-0.51	035.31	0
035	Sodium (%)	0169	0.3350	0.0100	0.3445	0.0186	0.0081	117	-0.51	035.32	0
035	Sodium (%)	0004	0.3350	0.0100	0.3445	0.0186	0.0081	117	-0.51	035.41	0
035	Sodium (%)	0011	0.3350	0.0130	0.3445	0.0186	0.0081	117	-0.51	035.41	0
035	Sodium (%)	0144	0.3350	0.0100	0.3445	0.0186	0.0081	117	-0.51	035.41	0
035	Sodium (%)	0354	0.3350	0.0100	0.3445	0.0186	0.0081	117	-0.51	035.41	0
035	Sodium (%)	0511	0.3350	0.0100	0.3445	0.0186	0.0081	117	-0.51	035.41	0
035	Sodium (%)	2252	0.3350	0.0100	0.3445	0.0186	0.0081	117	-0.51	035.41	0
035	Sodium (%)	0045	0.3350	0.0120	0.3445	0.0186	0.0081	117	-0.51	035.42	0
035	Sodium (%)	0278	0.3350	0.0100	0.3445	0.0186	0.0081	117	-0.51	035.42	0
035	Sodium (%)	0205	0.3355	0.0050	0.3445	0.0186	0.0081	117	-0.48	035.42	0
035	Sodium (%)	2146	0.3355	0.0110	0.3445	0.0186	0.0081	117	-0.48	035.51	0
035	Sodium (%)	2055	0.3357	0.0167	0.3445	0.0186	0.0081	117	-0.47	035.43	0
035	Sodium (%)	0512	0.3364	0.0095	0.3445	0.0186	0.0081	117	-0.44	035.43	0
035	Sodium (%)	2246	0.3369	0.0002	0.3445	0.0186	0.0081	117	-0.41	035.31	0
035	Sodium (%)	0142	0.3380	0.0037	0.3445	0.0186	0.0081	117	-0.35	035.31	0
035	Sodium (%)	0229	0.3385	0.0010	0.3445	0.0186	0.0081	117	-0.32	035.41	0
035	Sodium (%)	0510	0.3385	0.0030	0.3445	0.0186	0.0081	117	-0.32	035.43	0
035	Sodium (%)	2022	0.3395	0.0030	0.3445	0.0186	0.0081	117	-0.27	035.31	0
035	Sodium (%)	2018	0.3400	0.0200	0.3445	0.0186	0.0081	117	-0.24	035.31	0
035	Sodium (%)	0035	0.3400	0.0200	0.3445	0.0186	0.0081	117	-0.24	035.43	0
035	Sodium (%)	0019	0.3400	0.0000	0.3445	0.0186	0.0081	117	-0.24	035.41	0
035	Sodium (%)	2219	0.3400	0.0000	0.3445	0.0186	0.0081	117	-0.24	035.41	0
035	Sodium (%)	2207	0.3400	0.0000	0.3445	0.0186	0.0081	117	-0.24	035.42	0
035	Sodium (%)	0202	0.3400	0.0060	0.3445	0.0186	0.0081	117	-0.24	035.43	0
035	Sodium (%)	0682	0.3400	0.0000	0.3445	0.0186	0.0081	117	-0.24	035.43	0
035	Sodium (%)	0242	0.3400	0.0000	0.3445	0.0186	0.0081	117	-0.24	035.99	0
035	Sodium (%)	0263	0.3405	0.0034	0.3445	0.0186	0.0081	117	-0.21	035.42	0
035	Sodium (%)	0598	0.3406	0.0003	0.3445	0.0186	0.0081	117	-0.21	035.41	0
035	Sodium (%)	0407	0.3434	0.0012	0.3445	0.0186	0.0081	117	-0.06	035.41	0
035	Sodium (%)	0187	0.3445	0.0003	0.3445	0.0186	0.0081	117	0.00	035.42	0
035	Sodium (%)	2222	0.3450	0.0100	0.3445	0.0186	0.0081	117	0.03	035.31	0
035	Sodium (%)	0065	0.3450	0.0100	0.3445	0.0186	0.0081	117	0.03	035.41	0
035	Sodium (%)	0035	0.3450	0.0100	0.3445	0.0186	0.0081	117	0.03	035.42	0
035	Sodium (%)	0098	0.3450	0.0100	0.3445	0.0186	0.0081	117	0.03	035.43	0
035	Sodium (%)	0918	0.3450	0.0100	0.3445	0.0186	0.0081	117	0.03	035.43	0
035	Sodium (%)	0939	0.3450	0.0100	0.3445	0.0186	0.0081	117	0.03	035.53	0
035	Sodium (%)	0083	0.3455	0.0050	0.3445	0.0186	0.0081	117	0.06	035.43	0
035	Sodium (%)	0017	0.3460	0.0020	0.3445	0.0186	0.0081	117	0.08	035.43	0
035	Sodium (%)	0164	0.3475	0.0050	0.3445	0.0186	0.0081	117	0.16	035.41	0
035	Sodium (%)	2146	0.3485	0.0070	0.3445	0.0186	0.0081	117	0.22	035.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
035	Sodium (%)	0208	0.3490	0.0120	0.3445	0.0186	0.0081	117	0.24	035.41	0
035	Sodium (%)	0878	0.3500	0.0000	0.3445	0.0186	0.0081	117	0.30	035.41	0
035	Sodium (%)	0294	0.3500	0.0000	0.3445	0.0186	0.0081	117	0.30	035.42	0
035	Sodium (%)	0010	0.3500	0.0000	0.3445	0.0186	0.0081	117	0.30	035.43	0
035	Sodium (%)	0033	0.3500	0.0200	0.3445	0.0186	0.0081	117	0.30	035.43	0
035	Sodium (%)	0560	0.3504	0.0040	0.3445	0.0186	0.0081	117	0.32	035.42	0
035	Sodium (%)	0881	0.3505	0.0010	0.3445	0.0186	0.0081	117	0.33	035.01	0
035	Sodium (%)	0921	0.3520	0.0100	0.3445	0.0186	0.0081	117	0.41	035.31	0
035	Sodium (%)	2156	0.3525	0.0010	0.3445	0.0186	0.0081	117	0.43	035.41	0
035	Sodium (%)	0619	0.3535	0.0010	0.3445	0.0186	0.0081	117	0.49	035.41	0
035	Sodium (%)	0955	0.3535	0.0050	0.3445	0.0186	0.0081	117	0.49	035.41	0
035	Sodium (%)	0653	0.3540	0.0020	0.3445	0.0186	0.0081	117	0.51	035.41	0
035	Sodium (%)	2187	0.3547	0.0061	0.3445	0.0186	0.0081	117	0.55	035.43	0
035	Sodium (%)	0868	0.3550	0.0100	0.3445	0.0186	0.0081	117	0.57	035.31	0
035	Sodium (%)	0098	0.3550	0.0100	0.3445	0.0186	0.0081	117	0.57	035.41	0
035	Sodium (%)	0358	0.3550	0.0100	0.3445	0.0186	0.0081	117	0.57	035.41	0
035	Sodium (%)	0910	0.3550	0.0100	0.3445	0.0186	0.0081	117	0.57	035.41	0
035	Sodium (%)	0265	0.3550	0.0100	0.3445	0.0186	0.0081	117	0.57	035.42	0
035	Sodium (%)	0208	0.3555	0.0030	0.3445	0.0186	0.0081	117	0.59	035.31	0
035	Sodium (%)	0051	0.3570	0.0033	0.3445	0.0186	0.0081	117	0.67	035.41	0
035	Sodium (%)	2216	0.3600	0.0000	0.3445	0.0186	0.0081	117	0.84	035.32	0
035	Sodium (%)	0405	0.3600	0.0000	0.3445	0.0186	0.0081	117	0.84	035.41	0
035	Sodium (%)	0613	0.3600	0.0200	0.3445	0.0186	0.0081	117	0.84	035.42	0
035	Sodium (%)	0089	0.3600	0.0000	0.3445	0.0186	0.0081	117	0.84	035.43	0
035	Sodium (%)	0529	0.3620	0.0220	0.3445	0.0186	0.0081	117	0.94	035.31	0
035	Sodium (%)	0726	0.3625	0.0029	0.3445	0.0186	0.0081	117	0.97	035.42	0
035	Sodium (%)	0948	0.3630	0.0100	0.3445	0.0186	0.0081	117	1.00	035.31	0
035	Sodium (%)	0049	0.3650	0.0300	0.3445	0.0186	0.0081	117	1.11	035.41	0
035	Sodium (%)	0139	0.3650	0.0000	0.3445	0.0186	0.0081	117	1.11	035.41	0
035	Sodium (%)	0226	0.3650	0.0100	0.3445	0.0186	0.0081	117	1.11	035.41	0
035	Sodium (%)	0190	0.3650	0.0100	0.3445	0.0186	0.0081	117	1.11	035.42	0
035	Sodium (%)	2033	0.3650	0.0100	0.3445	0.0186	0.0081	117	1.11	035.43	0
035	Sodium (%)	0100	0.3650	0.0100	0.3445	0.0186	0.0081	117	1.11	035.99	0
035	Sodium (%)	0520	0.3693	0.0008	0.3445	0.0186	0.0081	117	1.34	035.41	0
035	Sodium (%)	0964	0.3698	0.0335	0.3445	0.0186	0.0081	117	1.36	035.43	0
035	Sodium (%)	2124	0.3700	0.0000	0.3445	0.0186	0.0081	117	1.37	035.43	0
035	Sodium (%)	2115	0.3710	0.0067	0.3445	0.0186	0.0081	117	1.43	035.31	0
035	Sodium (%)	0042	0.3710	0.0140	0.3445	0.0186	0.0081	117	1.43	035.42	0
035	Sodium (%)	0186	0.3715	0.0030	0.3445	0.0186	0.0081	117	1.45	035.52	0
035	Sodium (%)	0656	0.3750	0.0100	0.3445	0.0186	0.0081	117	1.64	035.31	0
035	Sodium (%)	0610	0.3850	0.0100	0.3445	0.0186	0.0081	117	2.18	035.43	0
035	Sodium (%)	0186	0.3900	0.0040	0.3445	0.0186	0.0081	117	2.45	035.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
035	Sodium (%)	0353	0.4100	0.0000	0.3445	0.0186	0.0081	117	3.53	035.43	0
035	Sodium (%)	0874	0.4135	0.0230	0.3445	0.0186	0.0081	117	3.71	035.31	0
035	Sodium (%)	0108	0.4200	0.0200	0.3445	0.0186	0.0081	117	4.06	035.05	0
035	Sodium (%)	0038	0.4200	0.0200	0.3445	0.0186	0.0081	117	4.06	035.43	0
035	Sodium (%)	0689	0.4300	0.0000	0.3445	0.0186	0.0081	117	4.60	035.31	0
035	Sodium (%)	2126	0.3500	0.0600	0.3445	0.0186	0.0081	117	0.30	035.31	1
035	Sodium (%)	0889	0.0600	0.1000	0.3445	0.0186	0.0081	117	-15.30	035.99	2
036	Sulfur (%)	0870	0.2227	0.0029	0.2502	0.0171	0.0078	40	-1.61	036.42	0
036	Sulfur (%)	0613	0.2250	0.0300	0.2502	0.0171	0.0078	40	-1.47	036.42	0
036	Sulfur (%)	0910	0.2270	0.0091	0.2502	0.0171	0.0078	40	-1.36	036.52	0
036	Sulfur (%)	2129	0.2280	0.0040	0.2502	0.0171	0.0078	40	-1.30	036.42	0
036	Sulfur (%)	0407	0.2308	0.0032	0.2502	0.0171	0.0078	40	-1.13	036.42	0
036	Sulfur (%)	0693	0.2315	0.0050	0.2502	0.0171	0.0078	40	-1.09	036.42	0
036	Sulfur (%)	0229	0.2340	0.0060	0.2502	0.0171	0.0078	40	-0.95	036.04	0
036	Sulfur (%)	0505	0.2350	0.0100	0.2502	0.0171	0.0078	40	-0.89	036.43	0
036	Sulfur (%)	0171	0.2360	0.0020	0.2502	0.0171	0.0078	40	-0.83	036.42	0
036	Sulfur (%)	0045	0.2390	0.0100	0.2502	0.0171	0.0078	40	-0.66	036.42	0
036	Sulfur (%)	0345	0.2390	0.0040	0.2502	0.0171	0.0078	40	-0.66	036.43	0
036	Sulfur (%)	0366	0.2400	0.0000	0.2502	0.0171	0.0078	40	-0.60	036.42	0
036	Sulfur (%)	0035	0.2400	0.0000	0.2502	0.0171	0.0078	40	-0.60	036.43	0
036	Sulfur (%)	0186	0.2405	0.0110	0.2502	0.0171	0.0078	40	-0.57	036.52	0
036	Sulfur (%)	0083	0.2420	0.0020	0.2502	0.0171	0.0078	40	-0.48	036.43	0
036	Sulfur (%)	0668	0.2435	0.0250	0.2502	0.0171	0.0078	40	-0.39	036.43	0
036	Sulfur (%)	0035	0.2500	0.0200	0.2502	0.0171	0.0078	40	-0.01	036.42	0
036	Sulfur (%)	0265	0.2500	0.0000	0.2502	0.0171	0.0078	40	-0.01	036.42	0
036	Sulfur (%)	0278	0.2500	0.0200	0.2502	0.0171	0.0078	40	-0.01	036.42	0
036	Sulfur (%)	2207	0.2500	0.0000	0.2502	0.0171	0.0078	40	-0.01	036.42	0
036	Sulfur (%)	0242	0.2500	0.0000	0.2502	0.0171	0.0078	40	-0.01	036.99	0
036	Sulfur (%)	0889	0.2541	0.0009	0.2502	0.0171	0.0078	40	0.22	036.99	0
036	Sulfur (%)	0186	0.2545	0.0010	0.2502	0.0171	0.0078	40	0.25	036.42	0
036	Sulfur (%)	0164	0.2550	0.0100	0.2502	0.0171	0.0078	40	0.28	036.42	0
036	Sulfur (%)	0708	0.2555	0.0010	0.2502	0.0171	0.0078	40	0.31	036.42	0
036	Sulfur (%)	0202	0.2555	0.0110	0.2502	0.0171	0.0078	40	0.31	036.43	0
036	Sulfur (%)	0187	0.2575	0.0029	0.2502	0.0171	0.0078	40	0.42	036.42	0
036	Sulfur (%)	0098	0.2584	0.0087	0.2502	0.0171	0.0078	40	0.47	036.43	0
036	Sulfur (%)	0560	0.2600	0.0036	0.2502	0.0171	0.0078	40	0.57	036.42	0
036	Sulfur (%)	0169	0.2600	0.0000	0.2502	0.0171	0.0078	40	0.57	036.43	0
036	Sulfur (%)	0357	0.2611	0.0033	0.2502	0.0171	0.0078	40	0.63	036.42	0
036	Sulfur (%)	0918	0.2620	0.0020	0.2502	0.0171	0.0078	40	0.69	036.43	0
036	Sulfur (%)	0037	0.2625	0.0030	0.2502	0.0171	0.0078	40	0.72	036.42	0
036	Sulfur (%)	0226	0.2650	0.0100	0.2502	0.0171	0.0078	40	0.86	036.04	0
036	Sulfur (%)	0353	0.2650	0.0100	0.2502	0.0171	0.0078	40	0.86	036.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
036	Sulfur (%)	0190	0.2750	0.0100	0.2502	0.0171	0.0078	40	1.45	036.42	0
036	Sulfur (%)	2089	0.2750	0.0100	0.2502	0.0171	0.0078	40	1.45	036.43	0
036	Sulfur (%)	0510	0.2800	0.0000	0.2502	0.0171	0.0078	40	1.74	036.43	0
036	Sulfur (%)	0294	0.2900	0.0000	0.2502	0.0171	0.0078	40	2.32	036.42	0
036	Sulfur (%)	0032	0.3100	0.0600	0.2502	0.0171	0.0078	40	3.49	036.43	0
036	Sulfur (%)	0726	180.8	0.8000	0.2502	0.0171	0.0078	40	10539.66	036.42	2
037	Zinc (mg / kg (ppm))	0609	0.0650	0.0100	170.2	12.81	4.435	105	-13.29	037.32	0
037	Zinc (mg / kg (ppm))	0726	0.2478	0.0005	170.2	12.81	4.435	105	-13.27	037.42	0
037	Zinc (mg / kg (ppm))	2128	115.4	0.0000	170.2	12.81	4.435	105	-4.28	037.32	0
037	Zinc (mg / kg (ppm))	0870	127.2	6.900	170.2	12.81	4.435	105	-3.36	037.42	0
037	Zinc (mg / kg (ppm))	0653	136.8	4.100	170.2	12.81	4.435	105	-2.61	037.41	0
037	Zinc (mg / kg (ppm))	0047	138.3	2.900	170.2	12.81	4.435	105	-2.49	037.52	0
037	Zinc (mg / kg (ppm))	0848	146.5	2.930	170.2	12.81	4.435	105	-1.85	037.41	0
037	Zinc (mg / kg (ppm))	0695	147.5	0.7030	170.2	12.81	4.435	105	-1.77	037.41	0
037	Zinc (mg / kg (ppm))	0955	149.0	2.000	170.2	12.81	4.435	105	-1.66	037.44	0
037	Zinc (mg / kg (ppm))	0242	149.5	9.000	170.2	12.81	4.435	105	-1.62	037.99	0
037	Zinc (mg / kg (ppm))	2055	150.9	5.900	170.2	12.81	4.435	105	-1.51	037.43	0
037	Zinc (mg / kg (ppm))	2089	153.2	0.6300	170.2	12.81	4.435	105	-1.33	037.43	0
037	Zinc (mg / kg (ppm))	2129	154.5	1.700	170.2	12.81	4.435	105	-1.23	037.99	0
037	Zinc (mg / kg (ppm))	0674	154.8	0.0000	170.2	12.81	4.435	105	-1.20	037.31	0
037	Zinc (mg / kg (ppm))	0668	156.0	10.00	170.2	12.81	4.435	105	-1.11	037.43	0
037	Zinc (mg / kg (ppm))	0689	157.5	7.000	170.2	12.81	4.435	105	-0.99	037.31	0
037	Zinc (mg / kg (ppm))	2114	158.6	4.413	170.2	12.81	4.435	105	-0.90	037.44	0
037	Zinc (mg / kg (ppm))	0049	159.3	2.250	170.2	12.81	4.435	105	-0.85	037.41	0
037	Zinc (mg / kg (ppm))	2022	160.0	2.000	170.2	12.81	4.435	105	-0.80	037.31	0
037	Zinc (mg / kg (ppm))	0693	160.2	4.435	170.2	12.81	4.435	105	-0.78	037.42	0
037	Zinc (mg / kg (ppm))	0948	160.3	0.7900	170.2	12.81	4.435	105	-0.77	037.33	0
037	Zinc (mg / kg (ppm))	0590	160.5	3.000	170.2	12.81	4.435	105	-0.76	037.99	0
037	Zinc (mg / kg (ppm))	0619	161.0	2.000	170.2	12.81	4.435	105	-0.72	037.41	0
037	Zinc (mg / kg (ppm))	0164	162.0	12.00	170.2	12.81	4.435	105	-0.64	037.41	0
037	Zinc (mg / kg (ppm))	0366	162.0	4.000	170.2	12.81	4.435	105	-0.64	037.42	0
037	Zinc (mg / kg (ppm))	0017	162.5	36.28	170.2	12.81	4.435	105	-0.60	037.43	0
037	Zinc (mg / kg (ppm))	0345	162.7	1.300	170.2	12.81	4.435	105	-0.59	037.43	0
037	Zinc (mg / kg (ppm))	0878	163.0	4.000	170.2	12.81	4.435	105	-0.56	037.41	0
037	Zinc (mg / kg (ppm))	0278	163.3	4.200	170.2	12.81	4.435	105	-0.54	037.42	0
037	Zinc (mg / kg (ppm))	0354	163.6	0.6700	170.2	12.81	4.435	105	-0.51	037.41	0
037	Zinc (mg / kg (ppm))	0051	163.7	1.050	170.2	12.81	4.435	105	-0.51	037.41	0
037	Zinc (mg / kg (ppm))	0425	163.8	1.280	170.2	12.81	4.435	105	-0.50	037.43	0
037	Zinc (mg / kg (ppm))	0187	164.1	0.6400	170.2	12.81	4.435	105	-0.48	037.42	0
037	Zinc (mg / kg (ppm))	0520	164.5	3.000	170.2	12.81	4.435	105	-0.44	037.41	0
037	Zinc (mg / kg (ppm))	0045	164.5	5.000	170.2	12.81	4.435	105	-0.44	037.42	0
037	Zinc (mg / kg (ppm))	0510	164.5	3.000	170.2	12.81	4.435	105	-0.44	037.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
037	Zinc (mg / kg (ppm))	0066	164.8	5.760	170.2	12.81	4.435	105	-0.42	037.44	0
037	Zinc (mg / kg (ppm))	0610	165.0	2.000	170.2	12.81	4.435	105	-0.41	037.43	0
037	Zinc (mg / kg (ppm))	0529	165.6	11.20	170.2	12.81	4.435	105	-0.36	037.31	0
037	Zinc (mg / kg (ppm))	2012	166.9	1.490	170.2	12.81	4.435	105	-0.25	037.41	0
037	Zinc (mg / kg (ppm))	0186	168.0	2.000	170.2	12.81	4.435	105	-0.17	037.42	0
037	Zinc (mg / kg (ppm))	2053	168.0	0.0000	170.2	12.81	4.435	105	-0.17	037.42	0
037	Zinc (mg / kg (ppm))	0033	168.0	0.0000	170.2	12.81	4.435	105	-0.17	037.43	0
037	Zinc (mg / kg (ppm))	0536	168.1	1.750	170.2	12.81	4.435	105	-0.17	037.31	0
037	Zinc (mg / kg (ppm))	0186	168.5	9.000	170.2	12.81	4.435	105	-0.13	037.52	0
037	Zinc (mg / kg (ppm))	0889	168.5	1.376	170.2	12.81	4.435	105	-0.13	037.99	0
037	Zinc (mg / kg (ppm))	2188	168.6	2.700	170.2	12.81	4.435	105	-0.13	037.31	0
037	Zinc (mg / kg (ppm))	0263	168.9	1.430	170.2	12.81	4.435	105	-0.10	037.42	0
037	Zinc (mg / kg (ppm))	0505	169.4	7.300	170.2	12.81	4.435	105	-0.07	037.43	0
037	Zinc (mg / kg (ppm))	0874	169.5	3.000	170.2	12.81	4.435	105	-0.05	037.31	0
037	Zinc (mg / kg (ppm))	0083	169.5	1.000	170.2	12.81	4.435	105	-0.05	037.43	0
037	Zinc (mg / kg (ppm))	0918	169.7	3.970	170.2	12.81	4.435	105	-0.04	037.43	0
037	Zinc (mg / kg (ppm))	0914	169.8	5.090	170.2	12.81	4.435	105	-0.03	037.53	0
037	Zinc (mg / kg (ppm))	0001	170.0	3.200	170.2	12.81	4.435	105	-0.02	037.31	0
037	Zinc (mg / kg (ppm))	0010	170.0	0.0000	170.2	12.81	4.435	105	-0.02	037.33	0
037	Zinc (mg / kg (ppm))	0682	171.0	0.0000	170.2	12.81	4.435	105	0.06	037.43	0
037	Zinc (mg / kg (ppm))	0011	171.1	4.816	170.2	12.81	4.435	105	0.07	037.41	0
037	Zinc (mg / kg (ppm))	0202	171.8	3.641	170.2	12.81	4.435	105	0.13	037.43	0
037	Zinc (mg / kg (ppm))	0208	172.0	0.9000	170.2	12.81	4.435	105	0.14	037.41	0
037	Zinc (mg / kg (ppm))	0265	172.0	8.000	170.2	12.81	4.435	105	0.14	037.42	0
037	Zinc (mg / kg (ppm))	0939	172.0	2.000	170.2	12.81	4.435	105	0.14	037.53	0
037	Zinc (mg / kg (ppm))	2192	172.2	18.81	170.2	12.81	4.435	105	0.16	037.43	0
037	Zinc (mg / kg (ppm))	0294	172.5	5.700	170.2	12.81	4.435	105	0.18	037.42	0
037	Zinc (mg / kg (ppm))	0560	172.5	4.400	170.2	12.81	4.435	105	0.18	037.42	0
037	Zinc (mg / kg (ppm))	0171	173.5	6.800	170.2	12.81	4.435	105	0.26	037.41	0
037	Zinc (mg / kg (ppm))	2207	173.5	1.000	170.2	12.81	4.435	105	0.26	037.42	0
037	Zinc (mg / kg (ppm))	0613	174.0	13.30	170.2	12.81	4.435	105	0.29	037.42	0
037	Zinc (mg / kg (ppm))	2196	174.0	0.0000	170.2	12.81	4.435	105	0.30	037.31	0
037	Zinc (mg / kg (ppm))	0074	174.5	13.00	170.2	12.81	4.435	105	0.34	037.41	0
037	Zinc (mg / kg (ppm))	2033	175.5	5.000	170.2	12.81	4.435	105	0.41	037.43	0
037	Zinc (mg / kg (ppm))	0675	175.7	6.440	170.2	12.81	4.435	105	0.43	037.43	0
037	Zinc (mg / kg (ppm))	0598	176.2	0.3000	170.2	12.81	4.435	105	0.47	037.41	0
037	Zinc (mg / kg (ppm))	0868	176.5	3.040	170.2	12.81	4.435	105	0.49	037.31	0
037	Zinc (mg / kg (ppm))	0407	176.7	6.285	170.2	12.81	4.435	105	0.51	037.41	0
037	Zinc (mg / kg (ppm))	0019	176.9	8.000	170.2	12.81	4.435	105	0.52	037.41	0
037	Zinc (mg / kg (ppm))	0964	177.1	0.3910	170.2	12.81	4.435	105	0.54	037.43	0
037	Zinc (mg / kg (ppm))	0169	177.5	1.000	170.2	12.81	4.435	105	0.57	037.43	0
037	Zinc (mg / kg (ppm))	0572	177.5	1.000	170.2	12.81	4.435	105	0.57	037.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
037	Zinc (mg / kg (ppm))	0229	178.1	2.400	170.2	12.81	4.435	105	0.62	037.41	0
037	Zinc (mg / kg (ppm))	0098	178.8	7.100	170.2	12.81	4.435	105	0.67	037.41	0
037	Zinc (mg / kg (ppm))	2115	179.0	0.2800	170.2	12.81	4.435	105	0.69	037.31	0
037	Zinc (mg / kg (ppm))	0226	179.6	4.750	170.2	12.81	4.435	105	0.73	037.41	0
037	Zinc (mg / kg (ppm))	0358	180.4	6.010	170.2	12.81	4.435	105	0.79	037.41	0
037	Zinc (mg / kg (ppm))	0357	180.5	1.176	170.2	12.81	4.435	105	0.81	037.42	0
037	Zinc (mg / kg (ppm))	0190	180.7	0.5400	170.2	12.81	4.435	105	0.82	037.42	0
037	Zinc (mg / kg (ppm))	0098	181.4	12.10	170.2	12.81	4.435	105	0.87	037.43	0
037	Zinc (mg / kg (ppm))	0656	181.5	16.10	170.2	12.81	4.435	105	0.88	037.31	0
037	Zinc (mg / kg (ppm))	0353	181.5	9.000	170.2	12.81	4.435	105	0.88	037.43	0
037	Zinc (mg / kg (ppm))	0968	183.0	4.000	170.2	12.81	4.435	105	1.00	037.43	0
037	Zinc (mg / kg (ppm))	0208	183.5	3.000	170.2	12.81	4.435	105	1.04	037.31	0
037	Zinc (mg / kg (ppm))	0035	183.5	7.000	170.2	12.81	4.435	105	1.04	037.42	0
037	Zinc (mg / kg (ppm))	2124	184.4	1.400	170.2	12.81	4.435	105	1.11	037.43	0
037	Zinc (mg / kg (ppm))	0511	184.5	1.000	170.2	12.81	4.435	105	1.12	037.41	0
037	Zinc (mg / kg (ppm))	0035	186.5	11.00	170.2	12.81	4.435	105	1.27	037.43	0
037	Zinc (mg / kg (ppm))	0512	186.8	1.400	170.2	12.81	4.435	105	1.30	037.43	0
037	Zinc (mg / kg (ppm))	0910	189.0	6.000	170.2	12.81	4.435	105	1.47	037.41	0
037	Zinc (mg / kg (ppm))	0028	189.2	0.8500	170.2	12.81	4.435	105	1.48	037.43	0
037	Zinc (mg / kg (ppm))	0100	191.0	4.000	170.2	12.81	4.435	105	1.62	037.99	0
037	Zinc (mg / kg (ppm))	0042	193.5	23.00	170.2	12.81	4.435	105	1.82	037.42	0
037	Zinc (mg / kg (ppm))	0405	195.0	0.0000	170.2	12.81	4.435	105	1.94	037.41	0
037	Zinc (mg / kg (ppm))	0037	198.8	0.4000	170.2	12.81	4.435	105	2.23	037.42	0
037	Zinc (mg / kg (ppm))	2144	202.8	0.5800	170.2	12.81	4.435	105	2.55	037.31	0
037	Zinc (mg / kg (ppm))	0038	203.5	7.000	170.2	12.81	4.435	105	2.60	037.43	0
037	Zinc (mg / kg (ppm))	0004	204.4	4.720	170.2	12.81	4.435	105	2.67	037.41	0
037	Zinc (mg / kg (ppm))	2246	296.3	5.700	170.2	12.81	4.435	105	9.85	037.33	0
037	Zinc (mg / kg (ppm))	0003	222.0	92.00	170.2	12.81	4.435	105	4.05	037.41	1
037	Zinc (mg / kg (ppm))	0032	228.0	76.00	170.2	12.81	4.435	105	4.51	037.43	1
037	Zinc (mg / kg (ppm))	2178	487.5	67.00	170.2	12.81	4.435	105	24.78	037.33	2
038	Molybdenum (mg / kg (ppm))	0693	1.290	0.7800	1.982	0.3137	0.1896	18	-2.21	038.42	0
038	Molybdenum (mg / kg (ppm))	0265	1.500	0.3200	1.982	0.3137	0.1896	18	-1.54	038.42	0
038	Molybdenum (mg / kg (ppm))	0038	1.676	0.0500	1.982	0.3137	0.1896	18	-0.97	038.43	0
038	Molybdenum (mg / kg (ppm))	0353	1.783	0.2588	1.982	0.3137	0.1896	18	-0.63	038.43	0
038	Molybdenum (mg / kg (ppm))	0345	1.800	0.0200	1.982	0.3137	0.1896	18	-0.58	038.43	0
038	Molybdenum (mg / kg (ppm))	0171	1.850	0.1000	1.982	0.3137	0.1896	18	-0.42	038.41	0
038	Molybdenum (mg / kg (ppm))	2055	1.850	0.1000	1.982	0.3137	0.1896	18	-0.42	038.43	0
038	Molybdenum (mg / kg (ppm))	0910	1.865	0.0500	1.982	0.3137	0.1896	18	-0.37	038.52	0
038	Molybdenum (mg / kg (ppm))	0226	1.950	0.2400	1.982	0.3137	0.1896	18	-0.10	038.41	0
038	Molybdenum (mg / kg (ppm))	0186	2.020	0.0600	1.982	0.3137	0.1896	18	0.12	038.52	0
038	Molybdenum (mg / kg (ppm))	0407	2.036	0.0082	1.982	0.3137	0.1896	18	0.17	038.41	0
038	Molybdenum (mg / kg (ppm))	0045	2.105	0.4500	1.982	0.3137	0.1896	18	0.39	038.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
038	Molybdenum (mg / kg (ppm))	0164	2.130	0.0800	1.982	0.3137	0.1896	18	0.47	038.53	0
038	Molybdenum (mg / kg (ppm))	0278	2.170	0.1000	1.982	0.3137	0.1896	18	0.60	038.42	0
038	Molybdenum (mg / kg (ppm))	0510	2.300	0.0000	1.982	0.3137	0.1896	18	1.01	038.43	0
038	Molybdenum (mg / kg (ppm))	0918	2.320	0.0957	1.982	0.3137	0.1896	18	1.08	038.53	0
038	Molybdenum (mg / kg (ppm))	0169	2.340	0.6000	1.982	0.3137	0.1896	18	1.14	038.43	0
038	Molybdenum (mg / kg (ppm))	0083	2.850	0.1000	1.982	0.3137	0.1896	18	2.77	038.43	0
038	Molybdenum (mg / kg (ppm))	2053	< 0.05		1.982	0.3137	0.1896	18		038.42	5
040	Barium (mg / kg (ppm))	2055	8.350	0.1000				3		040.43	0
040	Barium (mg / kg (ppm))	0918	8.644	0.1726				3		040.53	0
040	Barium (mg / kg (ppm))	0560	9.195	0.0100				3		040.42	0
042	Chloride (%)	0948	0.4770	0.0060				2		042.00	0
042	Chloride (%)	2124	0.4800	0.0000				2		042.00	0
101	Choline Chloride (mg / kg (ppm))	0051	1,399	25.00				3		101.02	0
101	Choline Chloride (mg / kg (ppm))	0227	1,871	68.00				3		101.02	0
101	Choline Chloride (mg / kg (ppm))	0939	2,761	116.0				3		101.02	0
102	Niacin (mg / kg (ppm))	2058	18.78	1.190	42.70	26.16	1.604	5		102.02	0
102	Niacin (mg / kg (ppm))	0160	29.05	3.700	42.70	26.16	1.604	5		102.02	0
102	Niacin (mg / kg (ppm))	2124	35.64	2.160	42.70	26.16	1.604	5		102.02	0
102	Niacin (mg / kg (ppm))	0939	43.42	0.7700	42.70	26.16	1.604	5		102.02	0
102	Niacin (mg / kg (ppm))	0227	86.60	0.2000	42.70	26.16	1.604	5		102.01	0
103	Pantothenic Acid (mg / kg (ppm))	2058	15.28	0.0600	16.73	1.179	0.6775	4		103.02	0
103	Pantothenic Acid (mg / kg (ppm))	0227	16.60	0.8000	16.73	1.179	0.6775	4		103.01	0
103	Pantothenic Acid (mg / kg (ppm))	0939	16.91	0.7500	16.73	1.179	0.6775	4		103.02	0
103	Pantothenic Acid (mg / kg (ppm))	0160	18.15	1.100	16.73	1.179	0.6775	4		103.02	0
103	Pantothenic Acid (mg / kg (ppm))	2124	15.97	4.720	16.73	1.179	0.6775	4		103.02	1
104	Riboflavin (mg / kg (ppm))	2124	3.225	0.2500	5.973	3.084	0.5267	6	-0.89	104.03	0
104	Riboflavin (mg / kg (ppm))	0160	3.675	0.3300	5.973	3.084	0.5267	6	-0.75	104.99	0
104	Riboflavin (mg / kg (ppm))	0910	5.010	0.2600	5.973	3.084	0.5267	6	-0.31	104.03	0
104	Riboflavin (mg / kg (ppm))	2058	5.345	0.0300	5.973	3.084	0.5267	6	-0.20	104.03	0
104	Riboflavin (mg / kg (ppm))	0227	8.660	0.8800	5.973	3.084	0.5267	6	0.87	104.00	0
104	Riboflavin (mg / kg (ppm))	0939	9.925	1.410	5.973	3.084	0.5267	6	1.28	104.03	0
105	Thiamine (mg / kg (ppm))	2058	4.590	0.2000	6.995	2.252	0.2133	6	-1.07	105.00	0
105	Thiamine (mg / kg (ppm))	0910	5.725	0.2700	6.995	2.252	0.2133	6	-0.56	105.00	0
105	Thiamine (mg / kg (ppm))	0939	6.370	0.3000	6.995	2.252	0.2133	6	-0.28	105.00	0
105	Thiamine (mg / kg (ppm))	0227	7.250	0.0400	6.995	2.252	0.2133	6	0.11	105.01	0
105	Thiamine (mg / kg (ppm))	2124	7.665	0.0700	6.995	2.252	0.2133	6	0.30	105.00	0
105	Thiamine (mg / kg (ppm))	0160	10.40	0.4000	6.995	2.252	0.2133	6	1.51	105.99	0
106	Vitamin A (KU / kg)	0169	2.100	0.2000	7.245	1.868	0.6131	19	-2.75	106.02	0
106	Vitamin A (KU / kg)	0905	4.249	0.2820	7.245	1.868	0.6131	19	-1.60	106.02	0
106	Vitamin A (KU / kg)	2053	5.150	0.1000	7.245	1.868	0.6131	19	-1.12	106.02	0
106	Vitamin A (KU / kg)	0227	5.860	0.5200	7.245	1.868	0.6131	19	-0.74	106.02	0
106	Vitamin A (KU / kg)	2192	6.287	0.2140	7.245	1.868	0.6131	19	-0.51	106.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
106	Vitamin A (KU / kg)	2103	6.735	0.4900	7.245	1.868	0.6131	19	-0.27	106.02	0
106	Vitamin A (KU / kg)	0910	6.750	0.1000	7.245	1.868	0.6131	19	-0.27	106.02	0
106	Vitamin A (KU / kg)	0004	7.130	1.940	7.245	1.868	0.6131	19	-0.06	106.02	0
106	Vitamin A (KU / kg)	0676	7.135	0.2100	7.245	1.868	0.6131	19	-0.06	106.02	0
106	Vitamin A (KU / kg)	0512	7.204	0.2910	7.245	1.868	0.6131	19	-0.02	106.02	0
106	Vitamin A (KU / kg)	0689	7.300	0.4000	7.245	1.868	0.6131	19	0.03	106.02	0
106	Vitamin A (KU / kg)	0610	7.450	0.1000	7.245	1.868	0.6131	19	0.11	106.02	0
106	Vitamin A (KU / kg)	0208	7.772	0.5510	7.245	1.868	0.6131	19	0.28	106.02	0
106	Vitamin A (KU / kg)	0010	7.915	0.9700	7.245	1.868	0.6131	19	0.36	106.02	0
106	Vitamin A (KU / kg)	2058	7.965	0.0100	7.245	1.868	0.6131	19	0.39	106.02	0
106	Vitamin A (KU / kg)	0098	7.980	0.3000	7.245	1.868	0.6131	19	0.39	106.01	0
106	Vitamin A (KU / kg)	0017	11.79	1.000	7.245	1.868	0.6131	19	2.43	106.02	0
106	Vitamin A (KU / kg)	0638	13.10	0.0000	7.245	1.868	0.6131	19	3.13	106.02	0
106	Vitamin A (KU / kg)	0675	179.2	3.970	7.245	1.868	0.6131	19	92.06	106.02	0
106	Vitamin A (KU / kg)	0171	8,205	1,090	7.245	1.868	0.6131	19	4388.08	106.00	2
106	Vitamin A (KU / kg)	0038	< 50		7.245	1.868	0.6131	19		106.02	5
107	Vitamin B12 (µg / kg (ppb))	0227	27.40	0.2000				2		107.00	0
107	Vitamin B12 (µg / kg (ppb))	0160	71.50	3.000				2		107.99	0
108	Vitamin D3 (KU / kg)	0227	0.6430	0.0440	1.096	0.6169	0.1420	7	-0.73	108.02	0
108	Vitamin D3 (KU / kg)	0051	0.6965	0.0130	1.096	0.6169	0.1420	7	-0.65	108.02	0
108	Vitamin D3 (KU / kg)	2058	0.7300	0.0000	1.096	0.6169	0.1420	7	-0.59	108.02	0
108	Vitamin D3 (KU / kg)	0208	0.7655	0.5090	1.096	0.6169	0.1420	7	-0.54	108.02	0
108	Vitamin D3 (KU / kg)	0910	1.165	0.0900	1.096	0.6169	0.1420	7	0.11	108.02	0
108	Vitamin D3 (KU / kg)	0169	1.650	0.1000	1.096	0.6169	0.1420	7	0.90	108.01	0
108	Vitamin D3 (KU / kg)	2192	3.182	0.2380	1.096	0.6169	0.1420	7	3.38	108.02	0
108	Vitamin D3 (KU / kg)	0675	243.8	3.390	1.096	0.6169	0.1420	7	393.39	108.02	1
108	Vitamin D3 (KU / kg)	0610	< 5000		1.096	0.6169	0.1420	7		108.02	5
109	Vitamin E (IU / kg)	0169	16.85	1.700	49.68	6.566	2.517	16	-5.00	109.02	0
109	Vitamin E (IU / kg)	2053	24.50	5.000	49.68	6.566	2.517	16	-3.83	109.02	0
109	Vitamin E (IU / kg)	0208	42.46	3.098	49.68	6.566	2.517	16	-1.10	109.02	0
109	Vitamin E (IU / kg)	0098	45.60	0.8000	49.68	6.566	2.517	16	-0.62	109.02	0
109	Vitamin E (IU / kg)	0638	47.30	0.0000	49.68	6.566	2.517	16	-0.36	109.02	0
109	Vitamin E (IU / kg)	0941	48.31	0.0000	49.68	6.566	2.517	16	-0.21	109.02	0
109	Vitamin E (IU / kg)	0610	49.00	4.000	49.68	6.566	2.517	16	-0.10	109.02	0
109	Vitamin E (IU / kg)	0910	49.30	0.4000	49.68	6.566	2.517	16	-0.06	109.02	0
109	Vitamin E (IU / kg)	0096	51.08	2.761	49.68	6.566	2.517	16	0.21	109.99	0
109	Vitamin E (IU / kg)	2103	51.50	7.000	49.68	6.566	2.517	16	0.28	109.02	0
109	Vitamin E (IU / kg)	0676	51.53	0.9800	49.68	6.566	2.517	16	0.28	109.02	0
109	Vitamin E (IU / kg)	2058	51.92	0.1100	49.68	6.566	2.517	16	0.34	109.02	0
109	Vitamin E (IU / kg)	0227	54.10	4.400	49.68	6.566	2.517	16	0.67	109.02	0
109	Vitamin E (IU / kg)	0675	56.04	3.640	49.68	6.566	2.517	16	0.97	109.02	0
109	Vitamin E (IU / kg)	0905	57.50	5.000	49.68	6.566	2.517	16	1.19	109.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
109	Vitamin E (IU / kg)	2192	60.35	1.380	49.68	6.566	2.517	16	1.63	109.02	0
109	Vitamin E (IU / kg)	0038	53.28	19.96	49.68	6.566	2.517	16	0.55	109.02	1
112	Pyridoxine (µg / g)	0160	1.050	0.0000	3.171	2.047	0.0775	4		112.01	0
112	Pyridoxine (µg / g)	2058	1.785	0.0300	3.171	2.047	0.0775	4		112.01	0
112	Pyridoxine (µg / g)	0227	4.890	0.1200	3.171	2.047	0.0775	4		112.01	0
112	Pyridoxine (µg / g)	0939	4.960	0.1600	3.171	2.047	0.0775	4		112.01	0
112	Pyridoxine (µg / g)	2124	4.500	1.000	3.171	2.047	0.0775	4		112.01	1
113	Folic Acid (mg / kg (ppm))	0939	0.3150	0.0500				3		113.02	0
113	Folic Acid (mg / kg (ppm))	0160	0.8700	0.0800				3		113.02	0
113	Folic Acid (mg / kg (ppm))	0227	0.9320	0.0180				3		113.01	0
114	Biotin (mg / kg (ppm))	0160	0.2550	0.0100				2		114.99	0
114	Biotin (mg / kg (ppm))	0227	0.3235	0.0590				2		114.01	0
118	Peroxide value (meq/kg)	2081	9.660	0.6200				1		118.99	0
120	Alanine (%)	0407	1.305	0.0300	1.490	0.0617	0.0201	37	-2.99	120.05	0
120	Alanine (%)	2022	1.355	0.0100	1.490	0.0617	0.0201	37	-2.18	120.00	0
120	Alanine (%)	0353	1.370	0.0200	1.490	0.0617	0.0201	37	-1.94	120.00	0
120	Alanine (%)	0160	1.385	0.0100	1.490	0.0617	0.0201	37	-1.70	120.99	0
120	Alanine (%)	2256	1.409	0.0030	1.490	0.0617	0.0201	37	-1.32	120.00	0
120	Alanine (%)	2144	1.410	0.0200	1.490	0.0617	0.0201	37	-1.29	120.99	0
120	Alanine (%)	2115	1.418	0.0310	1.490	0.0617	0.0201	37	-1.17	120.99	0
120	Alanine (%)	2188	1.445	0.0100	1.490	0.0617	0.0201	37	-0.72	120.05	0
120	Alanine (%)	0889	1.454	0.0366	1.490	0.0617	0.0201	37	-0.59	120.99	0
120	Alanine (%)	0675	1.460	0.0000	1.490	0.0617	0.0201	37	-0.48	120.00	0
120	Alanine (%)	0968	1.467	0.0490	1.490	0.0617	0.0201	37	-0.38	120.00	0
120	Alanine (%)	0910	1.480	0.0400	1.490	0.0617	0.0201	37	-0.16	120.00	0
120	Alanine (%)	0868	1.483	0.0090	1.490	0.0617	0.0201	37	-0.12	120.00	0
120	Alanine (%)	0878	1.485	0.0090	1.490	0.0617	0.0201	37	-0.08	120.00	0
120	Alanine (%)	0652	1.485	0.0300	1.490	0.0617	0.0201	37	-0.08	120.00	0
120	Alanine (%)	0610	1.490	0.0200	1.490	0.0617	0.0201	37	0.00	120.05	0
120	Alanine (%)	0859	1.490	0.0279	1.490	0.0617	0.0201	37	0.01	120.00	0
120	Alanine (%)	2059	1.491	0.0040	1.490	0.0617	0.0201	37	0.02	120.00	0
120	Alanine (%)	0918	1.492	0.0614	1.490	0.0617	0.0201	37	0.03	120.00	0
120	Alanine (%)	0619	1.495	0.0100	1.490	0.0617	0.0201	37	0.09	120.00	0
120	Alanine (%)	0939	1.495	0.0100	1.490	0.0617	0.0201	37	0.09	120.00	0
120	Alanine (%)	0644	1.501	0.0010	1.490	0.0617	0.0201	37	0.17	120.00	0
120	Alanine (%)	0504	1.505	0.0100	1.490	0.0617	0.0201	37	0.25	120.00	0
120	Alanine (%)	0676	1.505	0.0300	1.490	0.0617	0.0201	37	0.25	120.05	0
120	Alanine (%)	0354	1.510	0.0090	1.490	0.0617	0.0201	37	0.32	120.00	0
120	Alanine (%)	2196	1.512	0.0000	1.490	0.0617	0.0201	37	0.36	120.05	0
120	Alanine (%)	0571	1.514	0.0180	1.490	0.0617	0.0201	37	0.39	120.00	0
120	Alanine (%)	0872	1.515	0.0160	1.490	0.0617	0.0201	37	0.41	120.00	0
120	Alanine (%)	0227	1.515	0.0100	1.490	0.0617	0.0201	37	0.41	120.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
120	Alanine (%)	0870	1.542	0.0169	1.490	0.0617	0.0201	37	0.85	120.00	0
120	Alanine (%)	0626	1.547	0.0780	1.490	0.0617	0.0201	37	0.93	120.05	0
120	Alanine (%)	0098	1.549	0.0340	1.490	0.0617	0.0201	37	0.96	120.02	0
120	Alanine (%)	0684	1.555	0.0100	1.490	0.0617	0.0201	37	1.06	120.00	0
120	Alanine (%)	0682	1.569	0.0000	1.490	0.0617	0.0201	37	1.29	120.00	0
120	Alanine (%)	0171	1.599	0.0400	1.490	0.0617	0.0201	37	1.77	120.00	0
120	Alanine (%)	2146	1.625	0.0100	1.490	0.0617	0.0201	37	2.19	120.99	0
120	Alanine (%)	0226	1.673	0.0184	1.490	0.0617	0.0201	37	2.97	120.00	0
120	Alanine (%)	2246	1.376	0.1386	1.490	0.0617	0.0201	37	-1.85	120.05	1
121	Arginine (%)	0675	0.9150	0.0100	1.489	0.0585	0.0297	34	-9.82	121.00	0
121	Arginine (%)	0407	1.136	0.0520	1.489	0.0585	0.0297	34	-6.04	121.05	0
121	Arginine (%)	0889	1.243	0.0367	1.489	0.0585	0.0297	34	-4.22	121.99	0
121	Arginine (%)	2256	1.406	0.0160	1.489	0.0585	0.0297	34	-1.43	121.00	0
121	Arginine (%)	0682	1.417	0.0000	1.489	0.0585	0.0297	34	-1.24	121.00	0
121	Arginine (%)	0918	1.424	0.1524	1.489	0.0585	0.0297	34	-1.12	121.00	0
121	Arginine (%)	0160	1.455	0.1100	1.489	0.0585	0.0297	34	-0.59	121.99	0
121	Arginine (%)	0353	1.460	0.0000	1.489	0.0585	0.0297	34	-0.50	121.00	0
121	Arginine (%)	0939	1.465	0.0100	1.489	0.0585	0.0297	34	-0.42	121.00	0
121	Arginine (%)	2022	1.465	0.0100	1.489	0.0585	0.0297	34	-0.42	121.00	0
121	Arginine (%)	0870	1.472	0.0127	1.489	0.0585	0.0297	34	-0.30	121.00	0
121	Arginine (%)	0354	1.472	0.0020	1.489	0.0585	0.0297	34	-0.30	121.00	0
121	Arginine (%)	0868	1.476	0.0110	1.489	0.0585	0.0297	34	-0.24	121.00	0
121	Arginine (%)	2188	1.480	0.0000	1.489	0.0585	0.0297	34	-0.16	121.05	0
121	Arginine (%)	0968	1.481	0.0430	1.489	0.0585	0.0297	34	-0.15	121.00	0
121	Arginine (%)	2059	1.485	0.0030	1.489	0.0585	0.0297	34	-0.08	121.00	0
121	Arginine (%)	0227	1.485	0.0300	1.489	0.0585	0.0297	34	-0.07	121.00	0
121	Arginine (%)	0652	1.485	0.0300	1.489	0.0585	0.0297	34	-0.07	121.00	0
121	Arginine (%)	0684	1.490	0.0200	1.489	0.0585	0.0297	34	0.01	121.00	0
121	Arginine (%)	0226	1.493	0.0578	1.489	0.0585	0.0297	34	0.05	121.00	0
121	Arginine (%)	0859	1.494	0.0318	1.489	0.0585	0.0297	34	0.08	121.00	0
121	Arginine (%)	0504	1.495	0.0300	1.489	0.0585	0.0297	34	0.10	121.00	0
121	Arginine (%)	0098	1.501	0.0040	1.489	0.0585	0.0297	34	0.20	121.02	0
121	Arginine (%)	0619	1.520	0.0400	1.489	0.0585	0.0297	34	0.52	121.00	0
121	Arginine (%)	0878	1.523	0.0220	1.489	0.0585	0.0297	34	0.57	121.00	0
121	Arginine (%)	0872	1.528	0.0050	1.489	0.0585	0.0297	34	0.65	121.00	0
121	Arginine (%)	0910	1.535	0.0300	1.489	0.0585	0.0297	34	0.78	121.00	0
121	Arginine (%)	0644	1.537	0.0030	1.489	0.0585	0.0297	34	0.81	121.00	0
121	Arginine (%)	0571	1.540	0.0000	1.489	0.0585	0.0297	34	0.87	121.00	0
121	Arginine (%)	2115	1.560	0.0240	1.489	0.0585	0.0297	34	1.21	121.99	0
121	Arginine (%)	0171	1.565	0.1050	1.489	0.0585	0.0297	34	1.28	121.00	0
121	Arginine (%)	0610	1.575	0.0100	1.489	0.0585	0.0297	34	1.46	121.05	0
121	Arginine (%)	2196	1.603	0.0000	1.489	0.0585	0.0297	34	1.94	121.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
121	Arginine (%)	0676	1.640	0.1000	1.489	0.0585	0.0297	34	2.57	121.05	0
121	Arginine (%)	0626	1.576	0.2600	1.489	0.0585	0.0297	34	1.48	121.05	1
122	Aspartic (%)	0407	1.668	0.0010	1.932	0.0955	0.0307	36	-2.77	122.05	0
122	Aspartic (%)	0889	1.713	0.0046	1.932	0.0955	0.0307	36	-2.30	122.99	0
122	Aspartic (%)	2246	1.733	0.1297	1.932	0.0955	0.0307	36	-2.09	122.05	0
122	Aspartic (%)	0353	1.810	0.0000	1.932	0.0955	0.0307	36	-1.28	122.00	0
122	Aspartic (%)	0675	1.810	0.0200	1.932	0.0955	0.0307	36	-1.28	122.00	0
122	Aspartic (%)	2188	1.815	0.0300	1.932	0.0955	0.0307	36	-1.23	122.05	0
122	Aspartic (%)	2146	1.825	0.0100	1.932	0.0955	0.0307	36	-1.12	122.99	0
122	Aspartic (%)	2256	1.827	0.0070	1.932	0.0955	0.0307	36	-1.11	122.00	0
122	Aspartic (%)	2144	1.835	0.1100	1.932	0.0955	0.0307	36	-1.02	122.99	0
122	Aspartic (%)	0160	1.840	0.0000	1.932	0.0955	0.0307	36	-0.97	122.99	0
122	Aspartic (%)	0354	1.906	0.0260	1.932	0.0955	0.0307	36	-0.28	122.00	0
122	Aspartic (%)	0918	1.911	0.0549	1.932	0.0955	0.0307	36	-0.23	122.00	0
122	Aspartic (%)	0859	1.925	0.0501	1.932	0.0955	0.0307	36	-0.08	122.00	0
122	Aspartic (%)	0968	1.927	0.0090	1.932	0.0955	0.0307	36	-0.06	122.00	0
122	Aspartic (%)	0504	1.930	0.0200	1.932	0.0955	0.0307	36	-0.02	122.00	0
122	Aspartic (%)	2059	1.953	0.0020	1.932	0.0955	0.0307	36	0.22	122.00	0
122	Aspartic (%)	0652	1.955	0.0500	1.932	0.0955	0.0307	36	0.24	122.00	0
122	Aspartic (%)	0684	1.955	0.0100	1.932	0.0955	0.0307	36	0.24	122.00	0
122	Aspartic (%)	0682	1.956	0.0000	1.932	0.0955	0.0307	36	0.25	122.00	0
122	Aspartic (%)	0939	1.960	0.0000	1.932	0.0955	0.0307	36	0.29	122.00	0
122	Aspartic (%)	0571	1.962	0.0020	1.932	0.0955	0.0307	36	0.31	122.00	0
122	Aspartic (%)	0619	1.970	0.0000	1.932	0.0955	0.0307	36	0.39	122.00	0
122	Aspartic (%)	0878	1.975	0.0100	1.932	0.0955	0.0307	36	0.45	122.00	0
122	Aspartic (%)	2022	1.980	0.0000	1.932	0.0955	0.0307	36	0.50	122.00	0
122	Aspartic (%)	0227	1.985	0.0100	1.932	0.0955	0.0307	36	0.55	122.00	0
122	Aspartic (%)	0868	1.987	0.0210	1.932	0.0955	0.0307	36	0.57	122.00	0
122	Aspartic (%)	0872	1.988	0.0190	1.932	0.0955	0.0307	36	0.58	122.00	0
122	Aspartic (%)	0644	1.993	0.0010	1.932	0.0955	0.0307	36	0.63	122.00	0
122	Aspartic (%)	0171	2.000	0.1100	1.932	0.0955	0.0307	36	0.71	122.00	0
122	Aspartic (%)	0910	2.005	0.0500	1.932	0.0955	0.0307	36	0.76	122.00	0
122	Aspartic (%)	0676	2.005	0.1100	1.932	0.0955	0.0307	36	0.76	122.05	0
122	Aspartic (%)	0870	2.007	0.0027	1.932	0.0955	0.0307	36	0.78	122.00	0
122	Aspartic (%)	0610	2.015	0.0100	1.932	0.0955	0.0307	36	0.87	122.05	0
122	Aspartic (%)	0098	2.038	0.0190	1.932	0.0955	0.0307	36	1.10	122.02	0
122	Aspartic (%)	2196	2.077	0.0000	1.932	0.0955	0.0307	36	1.51	122.05	0
122	Aspartic (%)	0626	2.231	0.2070	1.932	0.0955	0.0307	36	3.12	122.05	0
122	Aspartic (%)	0226	2.346	0.6614	1.932	0.0955	0.0307	36	4.33	122.00	1
124	Cysteine/Cystine (%)	2246	0.0592	0.0137	0.3062	0.0310	0.0099	34	-7.97	124.05	0
124	Cysteine/Cystine (%)	0870	0.2556	0.0161	0.3062	0.0310	0.0099	34	-1.64	124.00	0
124	Cysteine/Cystine (%)	0171	0.2645	0.0090	0.3062	0.0310	0.0099	34	-1.35	124.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
124	Cysteine/Cystine (%)	2022	0.2700	0.0000	0.3062	0.0310	0.0099	34	-1.17	124.00	0
124	Cysteine/Cystine (%)	0227	0.2750	0.0100	0.3062	0.0310	0.0099	34	-1.01	124.00	0
124	Cysteine/Cystine (%)	0353	0.2800	0.0000	0.3062	0.0310	0.0099	34	-0.85	124.00	0
124	Cysteine/Cystine (%)	0644	0.2845	0.0070	0.3062	0.0310	0.0099	34	-0.70	124.00	0
124	Cysteine/Cystine (%)	0676	0.2850	0.0100	0.3062	0.0310	0.0099	34	-0.69	124.05	0
124	Cysteine/Cystine (%)	0652	0.2900	0.0000	0.3062	0.0310	0.0099	34	-0.52	124.00	0
124	Cysteine/Cystine (%)	0684	0.2900	0.0000	0.3062	0.0310	0.0099	34	-0.52	124.00	0
124	Cysteine/Cystine (%)	0939	0.2900	0.0000	0.3062	0.0310	0.0099	34	-0.52	124.00	0
124	Cysteine/Cystine (%)	2196	0.2900	0.0000	0.3062	0.0310	0.0099	34	-0.52	124.05	0
124	Cysteine/Cystine (%)	0859	0.2918	0.0083	0.3062	0.0310	0.0099	34	-0.47	124.00	0
124	Cysteine/Cystine (%)	0868	0.2980	0.0100	0.3062	0.0310	0.0099	34	-0.27	124.00	0
124	Cysteine/Cystine (%)	2188	0.3000	0.0000	0.3062	0.0310	0.0099	34	-0.20	124.05	0
124	Cysteine/Cystine (%)	0872	0.3030	0.0180	0.3062	0.0310	0.0099	34	-0.10	124.00	0
124	Cysteine/Cystine (%)	0682	0.3050	0.0000	0.3062	0.0310	0.0099	34	-0.04	124.00	0
124	Cysteine/Cystine (%)	0160	0.3050	0.0500	0.3062	0.0310	0.0099	34	-0.04	124.99	0
124	Cysteine/Cystine (%)	0878	0.3060	0.0020	0.3062	0.0310	0.0099	34	-0.01	124.00	0
124	Cysteine/Cystine (%)	0354	0.3100	0.0060	0.3062	0.0310	0.0099	34	0.12	124.00	0
124	Cysteine/Cystine (%)	0910	0.3100	0.0200	0.3062	0.0310	0.0099	34	0.12	124.00	0
124	Cysteine/Cystine (%)	0968	0.3100	0.0000	0.3062	0.0310	0.0099	34	0.12	124.00	0
124	Cysteine/Cystine (%)	0571	0.3120	0.0080	0.3062	0.0310	0.0099	34	0.19	124.00	0
124	Cysteine/Cystine (%)	2059	0.3160	0.0060	0.3062	0.0310	0.0099	34	0.31	124.00	0
124	Cysteine/Cystine (%)	0098	0.3230	0.0040	0.3062	0.0310	0.0099	34	0.54	124.02	0
124	Cysteine/Cystine (%)	0504	0.3250	0.0100	0.3062	0.0310	0.0099	34	0.61	124.00	0
124	Cysteine/Cystine (%)	0226	0.3294	0.0596	0.3062	0.0310	0.0099	34	0.75	124.00	0
124	Cysteine/Cystine (%)	2144	0.3350	0.0100	0.3062	0.0310	0.0099	34	0.93	124.99	0
124	Cysteine/Cystine (%)	2146	0.3400	0.0000	0.3062	0.0310	0.0099	34	1.09	124.99	0
124	Cysteine/Cystine (%)	0889	0.3462	0.0138	0.3062	0.0310	0.0099	34	1.29	124.99	0
124	Cysteine/Cystine (%)	0918	0.3500	0.0200	0.3062	0.0310	0.0099	34	1.41	124.00	0
124	Cysteine/Cystine (%)	0610	0.3600	0.0000	0.3062	0.0310	0.0099	34	1.73	124.05	0
124	Cysteine/Cystine (%)	0619	0.3770	0.0040	0.3062	0.0310	0.0099	34	2.28	124.00	0
124	Cysteine/Cystine (%)	0675	0.4100	0.0200	0.3062	0.0310	0.0099	34	3.35	124.00	0
124	Cysteine/Cystine (%)	0407	0.3325	0.0810	0.3062	0.0310	0.0099	34	0.85	124.05	1
125	Glutamic (%)	0407	3.252	0.0270	3.716	0.1909	0.0683	37	-2.44	125.05	0
125	Glutamic (%)	2144	3.270	0.0200	3.716	0.1909	0.0683	37	-2.34	125.99	0
125	Glutamic (%)	2188	3.325	0.1100	3.716	0.1909	0.0683	37	-2.05	125.05	0
125	Glutamic (%)	0675	3.450	0.0400	3.716	0.1909	0.0683	37	-1.40	125.00	0
125	Glutamic (%)	0171	3.478	0.1240	3.716	0.1909	0.0683	37	-1.25	125.00	0
125	Glutamic (%)	2256	3.512	0.0200	3.716	0.1909	0.0683	37	-1.07	125.00	0
125	Glutamic (%)	0353	3.520	0.0600	3.716	0.1909	0.0683	37	-1.03	125.00	0
125	Glutamic (%)	2022	3.555	0.0100	3.716	0.1909	0.0683	37	-0.85	125.00	0
125	Glutamic (%)	0160	3.575	0.0700	3.716	0.1909	0.0683	37	-0.74	125.99	0
125	Glutamic (%)	0652	3.620	0.0600	3.716	0.1909	0.0683	37	-0.51	125.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
125	Glutamic (%)	0968	3.628	0.1020	3.716	0.1909	0.0683	37	-0.46	125.00	0
125	Glutamic (%)	0354	3.667	0.0010	3.716	0.1909	0.0683	37	-0.26	125.00	0
125	Glutamic (%)	0676	3.670	0.1600	3.716	0.1909	0.0683	37	-0.24	125.05	0
125	Glutamic (%)	0859	3.671	0.1004	3.716	0.1909	0.0683	37	-0.24	125.00	0
125	Glutamic (%)	2196	3.706	0.0000	3.716	0.1909	0.0683	37	-0.05	125.05	0
125	Glutamic (%)	2059	3.714	0.0080	3.716	0.1909	0.0683	37	-0.01	125.00	0
125	Glutamic (%)	0504	3.715	0.0300	3.716	0.1909	0.0683	37	-0.01	125.00	0
125	Glutamic (%)	0571	3.716	0.0730	3.716	0.1909	0.0683	37	-0.01	125.00	0
125	Glutamic (%)	0868	3.728	0.0380	3.716	0.1909	0.0683	37	0.06	125.00	0
125	Glutamic (%)	0872	3.743	0.0560	3.716	0.1909	0.0683	37	0.14	125.00	0
125	Glutamic (%)	2246	3.748	0.2975	3.716	0.1909	0.0683	37	0.17	125.05	0
125	Glutamic (%)	0227	3.750	0.0200	3.716	0.1909	0.0683	37	0.18	125.00	0
125	Glutamic (%)	0610	3.750	0.0200	3.716	0.1909	0.0683	37	0.18	125.05	0
125	Glutamic (%)	0644	3.764	0.0090	3.716	0.1909	0.0683	37	0.25	125.00	0
125	Glutamic (%)	0889	3.775	0.0300	3.716	0.1909	0.0683	37	0.31	125.99	0
125	Glutamic (%)	0682	3.785	0.0000	3.716	0.1909	0.0683	37	0.36	125.00	0
125	Glutamic (%)	0878	3.796	0.0110	3.716	0.1909	0.0683	37	0.41	125.00	0
125	Glutamic (%)	0098	3.801	0.0520	3.716	0.1909	0.0683	37	0.44	125.02	0
125	Glutamic (%)	0684	3.805	0.1100	3.716	0.1909	0.0683	37	0.46	125.00	0
125	Glutamic (%)	0910	3.820	0.0400	3.716	0.1909	0.0683	37	0.54	125.00	0
125	Glutamic (%)	0939	3.895	0.0100	3.716	0.1909	0.0683	37	0.94	125.00	0
125	Glutamic (%)	0870	3.897	0.0118	3.716	0.1909	0.0683	37	0.95	125.00	0
125	Glutamic (%)	0619	3.955	0.0300	3.716	0.1909	0.0683	37	1.25	125.00	0
125	Glutamic (%)	0918	4.012	0.2891	3.716	0.1909	0.0683	37	1.55	125.00	0
125	Glutamic (%)	0226	4.186	0.1386	3.716	0.1909	0.0683	37	2.46	125.00	0
125	Glutamic (%)	2146	4.255	0.0300	3.716	0.1909	0.0683	37	2.82	125.99	0
125	Glutamic (%)	0626	4.373	0.3190	3.716	0.1909	0.0683	37	3.44	125.05	0
126	Glycine (%)	2146	1.535	0.0100	1.796	0.0708	0.0299	35	-3.69	126.99	0
126	Glycine (%)	2022	1.595	0.0100	1.796	0.0708	0.0299	35	-2.84	126.00	0
126	Glycine (%)	0407	1.634	0.0380	1.796	0.0708	0.0299	35	-2.29	126.05	0
126	Glycine (%)	0675	1.695	0.0300	1.796	0.0708	0.0299	35	-1.43	126.00	0
126	Glycine (%)	0353	1.705	0.0300	1.796	0.0708	0.0299	35	-1.29	126.00	0
126	Glycine (%)	2256	1.719	0.0060	1.796	0.0708	0.0299	35	-1.09	126.00	0
126	Glycine (%)	2188	1.725	0.0300	1.796	0.0708	0.0299	35	-1.00	126.05	0
126	Glycine (%)	0652	1.730	0.0200	1.796	0.0708	0.0299	35	-0.93	126.00	0
126	Glycine (%)	2246	1.731	0.1594	1.796	0.0708	0.0299	35	-0.92	126.05	0
126	Glycine (%)	2144	1.760	0.0600	1.796	0.0708	0.0299	35	-0.51	126.99	0
126	Glycine (%)	0354	1.764	0.0140	1.796	0.0708	0.0299	35	-0.45	126.00	0
126	Glycine (%)	2196	1.774	0.0000	1.796	0.0708	0.0299	35	-0.31	126.05	0
126	Glycine (%)	0859	1.789	0.0226	1.796	0.0708	0.0299	35	-0.10	126.00	0
126	Glycine (%)	0910	1.790	0.0200	1.796	0.0708	0.0299	35	-0.09	126.00	0
126	Glycine (%)	0504	1.800	0.0400	1.796	0.0708	0.0299	35	0.06	126.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
126	Glycine (%)	0571	1.801	0.0140	1.796	0.0708	0.0299	35	0.07	126.00	0
126	Glycine (%)	2059	1.806	0.0040	1.796	0.0708	0.0299	35	0.14	126.00	0
126	Glycine (%)	0878	1.808	0.0200	1.796	0.0708	0.0299	35	0.17	126.00	0
126	Glycine (%)	0939	1.810	0.0000	1.796	0.0708	0.0299	35	0.20	126.00	0
126	Glycine (%)	0872	1.816	0.0110	1.796	0.0708	0.0299	35	0.28	126.00	0
126	Glycine (%)	0682	1.822	0.0000	1.796	0.0708	0.0299	35	0.37	126.00	0
126	Glycine (%)	0870	1.822	0.0198	1.796	0.0708	0.0299	35	0.37	126.00	0
126	Glycine (%)	0968	1.824	0.0080	1.796	0.0708	0.0299	35	0.40	126.00	0
126	Glycine (%)	0098	1.825	0.0220	1.796	0.0708	0.0299	35	0.41	126.02	0
126	Glycine (%)	0619	1.830	0.0000	1.796	0.0708	0.0299	35	0.48	126.00	0
126	Glycine (%)	0644	1.834	0.0110	1.796	0.0708	0.0299	35	0.53	126.00	0
126	Glycine (%)	0227	1.835	0.0100	1.796	0.0708	0.0299	35	0.55	126.00	0
126	Glycine (%)	0684	1.845	0.0300	1.796	0.0708	0.0299	35	0.69	126.00	0
126	Glycine (%)	0918	1.851	0.1307	1.796	0.0708	0.0299	35	0.77	126.00	0
126	Glycine (%)	0868	1.859	0.0030	1.796	0.0708	0.0299	35	0.88	126.00	0
126	Glycine (%)	0610	1.865	0.0100	1.796	0.0708	0.0299	35	0.97	126.05	0
126	Glycine (%)	0171	1.867	0.1360	1.796	0.0708	0.0299	35	1.00	126.00	0
126	Glycine (%)	0676	1.890	0.0600	1.796	0.0708	0.0299	35	1.33	126.05	0
126	Glycine (%)	0226	1.899	0.0109	1.796	0.0708	0.0299	35	1.45	126.00	0
126	Glycine (%)	0626	2.068	0.0570	1.796	0.0708	0.0299	35	3.84	126.05	0
126	Glycine (%)	0160	1.640	0.1800	1.796	0.0708	0.0299	35	-2.20	126.99	1
126	Glycine (%)	0889	0.4200	0.0000	1.796	0.0708	0.0299	35	-19.44	126.99	2
127	Histidine (%)	0889	0.4200	0.0000	0.5780	0.0455	0.0099	37	-3.47	127.99	0
127	Histidine (%)	0675	0.4800	0.0200	0.5780	0.0455	0.0099	37	-2.15	127.00	0
127	Histidine (%)	0160	0.4800	0.0000	0.5780	0.0455	0.0099	37	-2.15	127.99	0
127	Histidine (%)	0407	0.4925	0.0470	0.5780	0.0455	0.0099	37	-1.88	127.05	0
127	Histidine (%)	2256	0.5210	0.0040	0.5780	0.0455	0.0099	37	-1.25	127.00	0
127	Histidine (%)	2146	0.5250	0.0100	0.5780	0.0455	0.0099	37	-1.17	127.99	0
127	Histidine (%)	2188	0.5350	0.0100	0.5780	0.0455	0.0099	37	-0.95	127.05	0
127	Histidine (%)	0968	0.5420	0.0260	0.5780	0.0455	0.0099	37	-0.79	127.00	0
127	Histidine (%)	0353	0.5450	0.0100	0.5780	0.0455	0.0099	37	-0.73	127.00	0
127	Histidine (%)	0918	0.5532	0.0129	0.5780	0.0455	0.0099	37	-0.55	127.00	0
127	Histidine (%)	0354	0.5550	0.0040	0.5780	0.0455	0.0099	37	-0.51	127.00	0
127	Histidine (%)	0939	0.5550	0.0100	0.5780	0.0455	0.0099	37	-0.51	127.00	0
127	Histidine (%)	0227	0.5700	0.0000	0.5780	0.0455	0.0099	37	-0.18	127.00	0
127	Histidine (%)	0571	0.5700	0.0020	0.5780	0.0455	0.0099	37	-0.18	127.00	0
127	Histidine (%)	0684	0.5750	0.0100	0.5780	0.0455	0.0099	37	-0.07	127.00	0
127	Histidine (%)	0098	0.5750	0.0060	0.5780	0.0455	0.0099	37	-0.07	127.02	0
127	Histidine (%)	2196	0.5780	0.0000	0.5780	0.0455	0.0099	37	0.00	127.05	0
127	Histidine (%)	2059	0.5785	0.0050	0.5780	0.0455	0.0099	37	0.01	127.00	0
127	Histidine (%)	0504	0.5800	0.0000	0.5780	0.0455	0.0099	37	0.04	127.00	0
127	Histidine (%)	0676	0.5800	0.0000	0.5780	0.0455	0.0099	37	0.04	127.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
127	Histidine (%)	0910	0.5850	0.0100	0.5780	0.0455	0.0099	37	0.15	127.00	0
127	Histidine (%)	0870	0.5883	0.0169	0.5780	0.0455	0.0099	37	0.22	127.00	0
127	Histidine (%)	0644	0.5900	0.0000	0.5780	0.0455	0.0099	37	0.26	127.00	0
127	Histidine (%)	0878	0.5900	0.0020	0.5780	0.0455	0.0099	37	0.26	127.00	0
127	Histidine (%)	0226	0.5913	0.0182	0.5780	0.0455	0.0099	37	0.29	127.00	0
127	Histidine (%)	0872	0.5940	0.0060	0.5780	0.0455	0.0099	37	0.35	127.00	0
127	Histidine (%)	0626	0.5950	0.0220	0.5780	0.0455	0.0099	37	0.37	127.05	0
127	Histidine (%)	2144	0.6050	0.0100	0.5780	0.0455	0.0099	37	0.59	127.99	0
127	Histidine (%)	0652	0.6200	0.0200	0.5780	0.0455	0.0099	37	0.92	127.00	0
127	Histidine (%)	0610	0.6200	0.0000	0.5780	0.0455	0.0099	37	0.92	127.05	0
127	Histidine (%)	0171	0.6225	0.0170	0.5780	0.0455	0.0099	37	0.98	127.00	0
127	Histidine (%)	2022	0.6250	0.0100	0.5780	0.0455	0.0099	37	1.03	127.00	0
127	Histidine (%)	0682	0.6270	0.0000	0.5780	0.0455	0.0099	37	1.08	127.00	0
127	Histidine (%)	0868	0.6305	0.0130	0.5780	0.0455	0.0099	37	1.15	127.00	0
127	Histidine (%)	0859	0.6339	0.0069	0.5780	0.0455	0.0099	37	1.23	127.00	0
127	Histidine (%)	0619	0.6995	0.0190	0.5780	0.0455	0.0099	37	2.67	127.00	0
127	Histidine (%)	2115	0.7755	0.0170	0.5780	0.0455	0.0099	37	4.34	127.99	0
127	Histidine (%)	2246	0.5369	0.0622	0.5780	0.0455	0.0099	37	-0.90	127.05	1
128	Isoleucine (%)	0353	0.6900	0.0200	0.8428	0.0558	0.0203	38	-2.74	128.00	0
128	Isoleucine (%)	0675	0.7000	0.0200	0.8428	0.0558	0.0203	38	-2.56	128.00	0
128	Isoleucine (%)	0407	0.7305	0.0750	0.8428	0.0558	0.0203	38	-2.01	128.05	0
128	Isoleucine (%)	2256	0.7460	0.0020	0.8428	0.0558	0.0203	38	-1.73	128.00	0
128	Isoleucine (%)	0160	0.7850	0.0300	0.8428	0.0558	0.0203	38	-1.04	128.99	0
128	Isoleucine (%)	0859	0.7852	0.0168	0.8428	0.0558	0.0203	38	-1.03	128.00	0
128	Isoleucine (%)	0968	0.7880	0.0080	0.8428	0.0558	0.0203	38	-0.98	128.00	0
128	Isoleucine (%)	0619	0.7915	0.0550	0.8428	0.0558	0.0203	38	-0.92	128.00	0
128	Isoleucine (%)	0354	0.7925	0.0030	0.8428	0.0558	0.0203	38	-0.90	128.00	0
128	Isoleucine (%)	2144	0.8000	0.0200	0.8428	0.0558	0.0203	38	-0.77	128.99	0
128	Isoleucine (%)	2246	0.8093	0.0844	0.8428	0.0558	0.0203	38	-0.60	128.05	0
128	Isoleucine (%)	0939	0.8100	0.0000	0.8428	0.0558	0.0203	38	-0.59	128.00	0
128	Isoleucine (%)	0868	0.8260	0.0180	0.8428	0.0558	0.0203	38	-0.30	128.00	0
128	Isoleucine (%)	0226	0.8305	0.0816	0.8428	0.0558	0.0203	38	-0.22	128.00	0
128	Isoleucine (%)	0872	0.8325	0.0070	0.8428	0.0558	0.0203	38	-0.19	128.00	0
128	Isoleucine (%)	0676	0.8350	0.0300	0.8428	0.0558	0.0203	38	-0.14	128.05	0
128	Isoleucine (%)	0684	0.8400	0.0200	0.8428	0.0558	0.0203	38	-0.05	128.00	0
128	Isoleucine (%)	0918	0.8421	0.0402	0.8428	0.0558	0.0203	38	-0.01	128.00	0
128	Isoleucine (%)	0652	0.8450	0.0100	0.8428	0.0558	0.0203	38	0.04	128.00	0
128	Isoleucine (%)	0889	0.8450	0.0100	0.8428	0.0558	0.0203	38	0.04	128.99	0
128	Isoleucine (%)	0227	0.8550	0.0100	0.8428	0.0558	0.0203	38	0.22	128.00	0
128	Isoleucine (%)	0682	0.8550	0.0000	0.8428	0.0558	0.0203	38	0.22	128.00	0
128	Isoleucine (%)	0171	0.8575	0.0450	0.8428	0.0558	0.0203	38	0.26	128.00	0
128	Isoleucine (%)	0571	0.8580	0.0120	0.8428	0.0558	0.0203	38	0.27	128.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
128	Isoleucine (%)	0878	0.8580	0.0080	0.8428	0.0558	0.0203	38	0.27	128.00	0
128	Isoleucine (%)	2188	0.8700	0.0000	0.8428	0.0558	0.0203	38	0.49	128.05	0
128	Isoleucine (%)	0910	0.8750	0.0100	0.8428	0.0558	0.0203	38	0.58	128.00	0
128	Isoleucine (%)	0870	0.8759	0.0036	0.8428	0.0558	0.0203	38	0.59	128.00	0
128	Isoleucine (%)	0098	0.8810	0.0060	0.8428	0.0558	0.0203	38	0.68	128.02	0
128	Isoleucine (%)	2059	0.8815	0.0070	0.8428	0.0558	0.0203	38	0.69	128.00	0
128	Isoleucine (%)	0644	0.8885	0.0110	0.8428	0.0558	0.0203	38	0.82	128.00	0
128	Isoleucine (%)	0610	0.8950	0.0300	0.8428	0.0558	0.0203	38	0.93	128.05	0
128	Isoleucine (%)	0626	0.9060	0.0240	0.8428	0.0558	0.0203	38	1.13	128.05	0
128	Isoleucine (%)	2115	0.9065	0.0230	0.8428	0.0558	0.0203	38	1.14	128.99	0
128	Isoleucine (%)	2146	0.9150	0.0100	0.8428	0.0558	0.0203	38	1.29	128.99	0
128	Isoleucine (%)	2196	0.9160	0.0000	0.8428	0.0558	0.0203	38	1.31	128.05	0
128	Isoleucine (%)	0504	0.9200	0.0000	0.8428	0.0558	0.0203	38	1.38	128.00	0
128	Isoleucine (%)	2022	0.9200	0.0200	0.8428	0.0558	0.0203	38	1.38	128.00	0
129	Leucine (%)	0407	1.624	0.0270	1.816	0.0538	0.0245	36	-3.58	129.05	0
129	Leucine (%)	0889	1.695	0.0100	1.816	0.0538	0.0245	36	-2.26	129.99	0
129	Leucine (%)	0675	1.710	0.0600	1.816	0.0538	0.0245	36	-1.98	129.00	0
129	Leucine (%)	0353	1.715	0.0300	1.816	0.0538	0.0245	36	-1.88	129.00	0
129	Leucine (%)	2188	1.740	0.0200	1.816	0.0538	0.0245	36	-1.42	129.05	0
129	Leucine (%)	2144	1.750	0.0000	1.816	0.0538	0.0245	36	-1.23	129.99	0
129	Leucine (%)	0859	1.755	0.0396	1.816	0.0538	0.0245	36	-1.15	129.00	0
129	Leucine (%)	0682	1.785	0.0000	1.816	0.0538	0.0245	36	-0.58	129.00	0
129	Leucine (%)	0652	1.785	0.0100	1.816	0.0538	0.0245	36	-0.58	129.00	0
129	Leucine (%)	2256	1.794	0.0110	1.816	0.0538	0.0245	36	-0.43	129.00	0
129	Leucine (%)	0939	1.795	0.0100	1.816	0.0538	0.0245	36	-0.40	129.00	0
129	Leucine (%)	2196	1.796	0.0000	1.816	0.0538	0.0245	36	-0.38	129.05	0
129	Leucine (%)	2115	1.804	0.0570	1.816	0.0538	0.0245	36	-0.24	129.99	0
129	Leucine (%)	0968	1.806	0.0790	1.816	0.0538	0.0245	36	-0.20	129.00	0
129	Leucine (%)	0676	1.810	0.0800	1.816	0.0538	0.0245	36	-0.12	129.05	0
129	Leucine (%)	0571	1.821	0.0120	1.816	0.0538	0.0245	36	0.08	129.00	0
129	Leucine (%)	0354	1.825	0.0160	1.816	0.0538	0.0245	36	0.16	129.00	0
129	Leucine (%)	0684	1.825	0.0100	1.816	0.0538	0.0245	36	0.16	129.00	0
129	Leucine (%)	0870	1.826	0.0034	1.816	0.0538	0.0245	36	0.18	129.00	0
129	Leucine (%)	0910	1.830	0.0400	1.816	0.0538	0.0245	36	0.25	129.00	0
129	Leucine (%)	2059	1.830	0.0020	1.816	0.0538	0.0245	36	0.25	129.00	0
129	Leucine (%)	0644	1.832	0.0170	1.816	0.0538	0.0245	36	0.28	129.00	0
129	Leucine (%)	2022	1.835	0.0100	1.816	0.0538	0.0245	36	0.35	129.00	0
129	Leucine (%)	0918	1.837	0.0495	1.816	0.0538	0.0245	36	0.37	129.00	0
129	Leucine (%)	0878	1.839	0.0130	1.816	0.0538	0.0245	36	0.41	129.00	0
129	Leucine (%)	0619	1.840	0.0400	1.816	0.0538	0.0245	36	0.44	129.00	0
129	Leucine (%)	0872	1.842	0.0270	1.816	0.0538	0.0245	36	0.47	129.00	0
129	Leucine (%)	0868	1.848	0.0230	1.816	0.0538	0.0245	36	0.58	129.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
129	Leucine (%)	0227	1.850	0.0000	1.816	0.0538	0.0245	36	0.62	129.00	0
129	Leucine (%)	0504	1.850	0.0000	1.816	0.0538	0.0245	36	0.62	129.00	0
129	Leucine (%)	0098	1.855	0.0210	1.816	0.0538	0.0245	36	0.71	129.02	0
129	Leucine (%)	0171	1.879	0.0880	1.816	0.0538	0.0245	36	1.16	129.00	0
129	Leucine (%)	0610	1.885	0.0100	1.816	0.0538	0.0245	36	1.27	129.05	0
129	Leucine (%)	2146	1.890	0.0000	1.816	0.0538	0.0245	36	1.37	129.99	0
129	Leucine (%)	0626	1.928	0.0190	1.816	0.0538	0.0245	36	2.06	129.05	0
129	Leucine (%)	0226	1.950	0.0467	1.816	0.0538	0.0245	36	2.48	129.00	0
129	Leucine (%)	0160	1.710	0.1600	1.816	0.0538	0.0245	36	-1.98	129.99	1
129	Leucine (%)	2246	1.859	0.1881	1.816	0.0538	0.0245	36	0.79	129.05	1
130	L-Lysine (%)	2053	1.009	0.0980	1.194	0.0895	0.0231	39	-2.06	130.99	0
130	L-Lysine (%)	0171	1.012	0.0720	1.194	0.0895	0.0231	39	-2.03	130.00	0
130	L-Lysine (%)	0407	1.015	0.0300	1.194	0.0895	0.0231	39	-2.00	130.05	0
130	L-Lysine (%)	0353	1.060	0.0000	1.194	0.0895	0.0231	39	-1.49	130.00	0
130	L-Lysine (%)	0848	1.100	0.0000	1.194	0.0895	0.0231	39	-1.05	130.00	0
130	L-Lysine (%)	0968	1.100	0.0120	1.194	0.0895	0.0231	39	-1.05	130.00	0
130	L-Lysine (%)	2188	1.105	0.0300	1.194	0.0895	0.0231	39	-0.99	130.05	0
130	L-Lysine (%)	2246	1.106	0.0526	1.194	0.0895	0.0231	39	-0.98	130.05	0
130	L-Lysine (%)	0918	1.129	0.0732	1.194	0.0895	0.0231	39	-0.72	130.00	0
130	L-Lysine (%)	0675	1.140	0.0200	1.194	0.0895	0.0231	39	-0.60	130.00	0
130	L-Lysine (%)	2256	1.146	0.0120	1.194	0.0895	0.0231	39	-0.53	130.00	0
130	L-Lysine (%)	2144	1.150	0.0200	1.194	0.0895	0.0231	39	-0.49	130.99	0
130	L-Lysine (%)	0652	1.155	0.0100	1.194	0.0895	0.0231	39	-0.43	130.00	0
130	L-Lysine (%)	0859	1.173	0.0290	1.194	0.0895	0.0231	39	-0.23	130.00	0
130	L-Lysine (%)	0939	1.175	0.0100	1.194	0.0895	0.0231	39	-0.21	130.00	0
130	L-Lysine (%)	0354	1.176	0.0130	1.194	0.0895	0.0231	39	-0.20	130.00	0
130	L-Lysine (%)	0872	1.178	0.0170	1.194	0.0895	0.0231	39	-0.18	130.00	0
130	L-Lysine (%)	0684	1.185	0.0100	1.194	0.0895	0.0231	39	-0.10	130.00	0
130	L-Lysine (%)	0619	1.205	0.0300	1.194	0.0895	0.0231	39	0.13	130.00	0
130	L-Lysine (%)	2059	1.206	0.0070	1.194	0.0895	0.0231	39	0.13	130.00	0
130	L-Lysine (%)	0868	1.207	0.0150	1.194	0.0895	0.0231	39	0.14	130.00	0
130	L-Lysine (%)	0682	1.208	0.0000	1.194	0.0895	0.0231	39	0.16	130.00	0
130	L-Lysine (%)	0571	1.209	0.0110	1.194	0.0895	0.0231	39	0.16	130.00	0
130	L-Lysine (%)	0878	1.210	0.0140	1.194	0.0895	0.0231	39	0.18	130.00	0
130	L-Lysine (%)	0910	1.210	0.0400	1.194	0.0895	0.0231	39	0.18	130.00	0
130	L-Lysine (%)	0676	1.215	0.0300	1.194	0.0895	0.0231	39	0.24	130.05	0
130	L-Lysine (%)	0644	1.226	0.0010	1.194	0.0895	0.0231	39	0.35	130.00	0
130	L-Lysine (%)	0610	1.230	0.0000	1.194	0.0895	0.0231	39	0.40	130.05	0
130	L-Lysine (%)	2196	1.231	0.0000	1.194	0.0895	0.0231	39	0.42	130.05	0
130	L-Lysine (%)	2022	1.240	0.0200	1.194	0.0895	0.0231	39	0.52	130.00	0
130	L-Lysine (%)	0504	1.250	0.0000	1.194	0.0895	0.0231	39	0.63	130.00	0
130	L-Lysine (%)	2115	1.273	0.0540	1.194	0.0895	0.0231	39	0.89	130.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
130	L-Lysine (%)	0226	1.286	0.0301	1.194	0.0895	0.0231	39	1.03	130.00	0
130	L-Lysine (%)	0626	1.291	0.0170	1.194	0.0895	0.0231	39	1.08	130.05	0
130	L-Lysine (%)	0870	1.295	0.0128	1.194	0.0895	0.0231	39	1.13	130.00	0
130	L-Lysine (%)	2146	1.330	0.0200	1.194	0.0895	0.0231	39	1.52	130.99	0
130	L-Lysine (%)	0889	1.340	0.0200	1.194	0.0895	0.0231	39	1.63	130.99	0
130	L-Lysine (%)	0098	1.343	0.0120	1.194	0.0895	0.0231	39	1.67	130.02	0
130	L-Lysine (%)	0227	1.350	0.0600	1.194	0.0895	0.0231	39	1.75	130.00	0
130	L-Lysine (%)	0160	1.135	0.1300	1.194	0.0895	0.0231	39	-0.66	130.99	1
131	Methionine (%)	2246	0.3119	0.0216	0.3892	0.0287	0.0080	37	-2.69	131.05	0
131	Methionine (%)	0682	0.3480	0.0000	0.3892	0.0287	0.0080	37	-1.43	131.00	0
131	Methionine (%)	0684	0.3500	0.0200	0.3892	0.0287	0.0080	37	-1.36	131.00	0
131	Methionine (%)	0848	0.3500	0.0000	0.3892	0.0287	0.0080	37	-1.36	131.00	0
131	Methionine (%)	0171	0.3535	0.0390	0.3892	0.0287	0.0080	37	-1.24	131.00	0
131	Methionine (%)	2196	0.3610	0.0000	0.3892	0.0287	0.0080	37	-0.98	131.05	0
131	Methionine (%)	0226	0.3629	0.0259	0.3892	0.0287	0.0080	37	-0.92	131.00	0
131	Methionine (%)	0870	0.3632	0.0194	0.3892	0.0287	0.0080	37	-0.90	131.00	0
131	Methionine (%)	2188	0.3650	0.0100	0.3892	0.0287	0.0080	37	-0.84	131.05	0
131	Methionine (%)	0968	0.3720	0.0000	0.3892	0.0287	0.0080	37	-0.60	131.00	0
131	Methionine (%)	0354	0.3775	0.0010	0.3892	0.0287	0.0080	37	-0.41	131.00	0
131	Methionine (%)	0859	0.3794	0.0080	0.3892	0.0287	0.0080	37	-0.34	131.00	0
131	Methionine (%)	0160	0.3800	0.0000	0.3892	0.0287	0.0080	37	-0.32	131.99	0
131	Methionine (%)	0227	0.3850	0.0100	0.3892	0.0287	0.0080	37	-0.14	131.00	0
131	Methionine (%)	0619	0.3865	0.0090	0.3892	0.0287	0.0080	37	-0.09	131.00	0
131	Methionine (%)	2115	0.3865	0.0170	0.3892	0.0287	0.0080	37	-0.09	131.99	0
131	Methionine (%)	0878	0.3895	0.0150	0.3892	0.0287	0.0080	37	0.01	131.00	0
131	Methionine (%)	0353	0.3900	0.0000	0.3892	0.0287	0.0080	37	0.03	131.00	0
131	Methionine (%)	0504	0.3900	0.0000	0.3892	0.0287	0.0080	37	0.03	131.00	0
131	Methionine (%)	0652	0.3900	0.0200	0.3892	0.0287	0.0080	37	0.03	131.00	0
131	Methionine (%)	0939	0.3900	0.0000	0.3892	0.0287	0.0080	37	0.03	131.00	0
131	Methionine (%)	2022	0.3900	0.0000	0.3892	0.0287	0.0080	37	0.03	131.00	0
131	Methionine (%)	0626	0.3910	0.0080	0.3892	0.0287	0.0080	37	0.06	131.05	0
131	Methionine (%)	0098	0.3925	0.0010	0.3892	0.0287	0.0080	37	0.12	131.02	0
131	Methionine (%)	2059	0.3990	0.0060	0.3892	0.0287	0.0080	37	0.34	131.00	0
131	Methionine (%)	0571	0.3995	0.0070	0.3892	0.0287	0.0080	37	0.36	131.00	0
131	Methionine (%)	0872	0.4045	0.0030	0.3892	0.0287	0.0080	37	0.53	131.00	0
131	Methionine (%)	0675	0.4050	0.0100	0.3892	0.0287	0.0080	37	0.55	131.00	0
131	Methionine (%)	0610	0.4050	0.0100	0.3892	0.0287	0.0080	37	0.55	131.05	0
131	Methionine (%)	0676	0.4050	0.0100	0.3892	0.0287	0.0080	37	0.55	131.05	0
131	Methionine (%)	0868	0.4095	0.0110	0.3892	0.0287	0.0080	37	0.71	131.00	0
131	Methionine (%)	0644	0.4250	0.0020	0.3892	0.0287	0.0080	37	1.25	131.00	0
131	Methionine (%)	0910	0.4300	0.0000	0.3892	0.0287	0.0080	37	1.42	131.00	0
131	Methionine (%)	0918	0.4300	0.0000	0.3892	0.0287	0.0080	37	1.42	131.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
131	Methionine (%)	0407	0.4412	0.0003	0.3892	0.0287	0.0080	37	1.81	131.05	0
131	Methionine (%)	2146	0.4650	0.0100	0.3892	0.0287	0.0080	37	2.64	131.99	0
131	Methionine (%)	0889	0.5300	0.0000	0.3892	0.0287	0.0080	37	4.91	131.99	0
131	Methionine (%)	2144	0.7750	0.0500	0.3892	0.0287	0.0080	37	13.44	131.99	2
132	Phenylalanine (%)	0407	0.8300	0.0260	0.9967	0.0538	0.0143	37	-3.10	132.05	0
132	Phenylalanine (%)	0889	0.8700	0.0000	0.9967	0.0538	0.0143	37	-2.36	132.99	0
132	Phenylalanine (%)	0160	0.9100	0.0000	0.9967	0.0538	0.0143	37	-1.61	132.99	0
132	Phenylalanine (%)	2256	0.9225	0.0030	0.9967	0.0538	0.0143	37	-1.38	132.00	0
132	Phenylalanine (%)	0675	0.9250	0.0100	0.9967	0.0538	0.0143	37	-1.33	132.00	0
132	Phenylalanine (%)	0918	0.9338	0.0181	0.9967	0.0538	0.0143	37	-1.17	132.00	0
132	Phenylalanine (%)	0968	0.9360	0.0140	0.9967	0.0538	0.0143	37	-1.13	132.00	0
132	Phenylalanine (%)	2188	0.9450	0.0100	0.9967	0.0538	0.0143	37	-0.96	132.05	0
132	Phenylalanine (%)	0939	0.9500	0.0200	0.9967	0.0538	0.0143	37	-0.87	132.00	0
132	Phenylalanine (%)	0684	0.9600	0.0000	0.9967	0.0538	0.0143	37	-0.68	132.00	0
132	Phenylalanine (%)	0682	0.9730	0.0000	0.9967	0.0538	0.0143	37	-0.44	132.00	0
132	Phenylalanine (%)	0868	0.9775	0.0230	0.9967	0.0538	0.0143	37	-0.36	132.00	0
132	Phenylalanine (%)	0227	0.9800	0.0200	0.9967	0.0538	0.0143	37	-0.31	132.00	0
132	Phenylalanine (%)	2059	0.9905	0.0110	0.9967	0.0538	0.0143	37	-0.11	132.00	0
132	Phenylalanine (%)	0870	0.9948	0.0330	0.9967	0.0538	0.0143	37	-0.03	132.00	0
132	Phenylalanine (%)	0878	0.9990	0.0020	0.9967	0.0538	0.0143	37	0.04	132.00	0
132	Phenylalanine (%)	0652	1.000	0.0200	0.9967	0.0538	0.0143	37	0.06	132.00	0
132	Phenylalanine (%)	2196	1.003	0.0000	0.9967	0.0538	0.0143	37	0.12	132.05	0
132	Phenylalanine (%)	0354	1.005	0.0110	0.9967	0.0538	0.0143	37	0.15	132.00	0
132	Phenylalanine (%)	0098	1.006	0.0150	0.9967	0.0538	0.0143	37	0.16	132.02	0
132	Phenylalanine (%)	0226	1.007	0.0079	0.9967	0.0538	0.0143	37	0.20	132.00	0
132	Phenylalanine (%)	0353	1.010	0.0200	0.9967	0.0538	0.0143	37	0.25	132.00	0
132	Phenylalanine (%)	0872	1.012	0.0130	0.9967	0.0538	0.0143	37	0.28	132.00	0
132	Phenylalanine (%)	0571	1.018	0.0010	0.9967	0.0538	0.0143	37	0.39	132.00	0
132	Phenylalanine (%)	0626	1.018	0.0060	0.9967	0.0538	0.0143	37	0.40	132.05	0
132	Phenylalanine (%)	0171	1.024	0.0530	0.9967	0.0538	0.0143	37	0.50	132.00	0
132	Phenylalanine (%)	0610	1.025	0.0100	0.9967	0.0538	0.0143	37	0.53	132.05	0
132	Phenylalanine (%)	0644	1.031	0.0010	0.9967	0.0538	0.0143	37	0.63	132.00	0
132	Phenylalanine (%)	2022	1.035	0.0100	0.9967	0.0538	0.0143	37	0.71	132.00	0
132	Phenylalanine (%)	0676	1.035	0.0300	0.9967	0.0538	0.0143	37	0.71	132.05	0
132	Phenylalanine (%)	2144	1.045	0.0300	0.9967	0.0538	0.0143	37	0.90	132.99	0
132	Phenylalanine (%)	2115	1.047	0.0180	0.9967	0.0538	0.0143	37	0.94	132.99	0
132	Phenylalanine (%)	0859	1.051	0.0236	0.9967	0.0538	0.0143	37	1.01	132.00	0
132	Phenylalanine (%)	0504	1.060	0.0000	0.9967	0.0538	0.0143	37	1.18	132.00	0
132	Phenylalanine (%)	2146	1.060	0.0000	0.9967	0.0538	0.0143	37	1.18	132.99	0
132	Phenylalanine (%)	0910	1.075	0.0300	0.9967	0.0538	0.0143	37	1.46	132.00	0
132	Phenylalanine (%)	0619	1.080	0.0400	0.9967	0.0538	0.0143	37	1.55	132.00	0
132	Phenylalanine (%)	2246	0.9416	0.0742	0.9967	0.0538	0.0143	37	-1.02	132.05	1

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
133	Proline (%)	2256	1.413	0.0130	1.587	0.0905	0.0357	35	-1.93	133.00	0
133	Proline (%)	2144	1.470	0.0800	1.587	0.0905	0.0357	35	-1.29	133.99	0
133	Proline (%)	0939	1.480	0.0200	1.587	0.0905	0.0357	35	-1.18	133.00	0
133	Proline (%)	0859	1.483	0.0070	1.587	0.0905	0.0357	35	-1.15	133.00	0
133	Proline (%)	0407	1.492	0.0350	1.587	0.0905	0.0357	35	-1.06	133.05	0
133	Proline (%)	0868	1.494	0.0150	1.587	0.0905	0.0357	35	-1.03	133.00	0
133	Proline (%)	0682	1.508	0.0000	1.587	0.0905	0.0357	35	-0.87	133.00	0
133	Proline (%)	0504	1.510	0.0400	1.587	0.0905	0.0357	35	-0.85	133.00	0
133	Proline (%)	0160	1.510	0.0200	1.587	0.0905	0.0357	35	-0.85	133.99	0
133	Proline (%)	0353	1.520	0.0000	1.587	0.0905	0.0357	35	-0.74	133.00	0
133	Proline (%)	0870	1.522	0.0190	1.587	0.0905	0.0357	35	-0.72	133.00	0
133	Proline (%)	0889	1.530	0.0400	1.587	0.0905	0.0357	35	-0.63	133.99	0
133	Proline (%)	0968	1.547	0.0330	1.587	0.0905	0.0357	35	-0.45	133.00	0
133	Proline (%)	0652	1.560	0.0200	1.587	0.0905	0.0357	35	-0.30	133.00	0
133	Proline (%)	0675	1.560	0.2000	1.587	0.0905	0.0357	35	-0.30	133.00	0
133	Proline (%)	0910	1.570	0.0800	1.587	0.0905	0.0357	35	-0.19	133.00	0
133	Proline (%)	0571	1.577	0.0170	1.587	0.0905	0.0357	35	-0.12	133.00	0
133	Proline (%)	0872	1.593	0.0370	1.587	0.0905	0.0357	35	0.06	133.00	0
133	Proline (%)	0354	1.597	0.0100	1.587	0.0905	0.0357	35	0.11	133.00	0
133	Proline (%)	0676	1.600	0.1200	1.587	0.0905	0.0357	35	0.14	133.05	0
133	Proline (%)	0878	1.608	0.0310	1.587	0.0905	0.0357	35	0.23	133.00	0
133	Proline (%)	0684	1.630	0.0400	1.587	0.0905	0.0357	35	0.47	133.00	0
133	Proline (%)	0610	1.635	0.0100	1.587	0.0905	0.0357	35	0.53	133.05	0
133	Proline (%)	2022	1.640	0.0000	1.587	0.0905	0.0357	35	0.58	133.00	0
133	Proline (%)	0227	1.640	0.0400	1.587	0.0905	0.0357	35	0.58	133.00	0
133	Proline (%)	0171	1.646	0.0110	1.587	0.0905	0.0357	35	0.65	133.00	0
133	Proline (%)	2196	1.648	0.0000	1.587	0.0905	0.0357	35	0.67	133.05	0
133	Proline (%)	2188	1.655	0.0700	1.587	0.0905	0.0357	35	0.75	133.05	0
133	Proline (%)	2059	1.658	0.0100	1.587	0.0905	0.0357	35	0.78	133.00	0
133	Proline (%)	2146	1.670	0.0000	1.587	0.0905	0.0357	35	0.92	133.99	0
133	Proline (%)	0644	1.675	0.0090	1.587	0.0905	0.0357	35	0.97	133.00	0
133	Proline (%)	0626	1.704	0.0310	1.587	0.0905	0.0357	35	1.29	133.05	0
133	Proline (%)	0918	1.731	0.0286	1.587	0.0905	0.0357	35	1.58	133.00	0
133	Proline (%)	0619	1.865	0.0300	1.587	0.0905	0.0357	35	3.07	133.00	0
133	Proline (%)	0226	2.181	0.1317	1.587	0.0905	0.0357	35	6.56	133.00	0
133	Proline (%)	2246	3.289	0.4257	1.587	0.0905	0.0357	35	18.80	133.05	2
134	Serine (%)	0504	0.8850	0.0100	1.006	0.0579	0.0179	35	-2.09	134.00	0
134	Serine (%)	0407	0.8950	0.0180	1.006	0.0579	0.0179	35	-1.92	134.05	0
134	Serine (%)	2188	0.9150	0.0100	1.006	0.0579	0.0179	35	-1.57	134.05	0
134	Serine (%)	0160	0.9150	0.0100	1.006	0.0579	0.0179	35	-1.57	134.99	0
134	Serine (%)	0098	0.9445	0.0270	1.006	0.0579	0.0179	35	-1.06	134.02	0
134	Serine (%)	0644	0.9565	0.0050	1.006	0.0579	0.0179	35	-0.86	134.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
134	Serine (%)	2256	0.9625	0.0070	1.006	0.0579	0.0179	35	-0.75	134.00	0
134	Serine (%)	2059	0.9630	0.0060	1.006	0.0579	0.0179	35	-0.74	134.00	0
134	Serine (%)	0353	0.9800	0.0000	1.006	0.0579	0.0179	35	-0.45	134.00	0
134	Serine (%)	0675	0.9800	0.0000	1.006	0.0579	0.0179	35	-0.45	134.00	0
134	Serine (%)	0610	0.9800	0.0000	1.006	0.0579	0.0179	35	-0.45	134.05	0
134	Serine (%)	0968	0.9805	0.0150	1.006	0.0579	0.0179	35	-0.44	134.00	0
134	Serine (%)	0652	0.9850	0.0300	1.006	0.0579	0.0179	35	-0.36	134.00	0
134	Serine (%)	2022	0.9950	0.0300	1.006	0.0579	0.0179	35	-0.19	134.00	0
134	Serine (%)	0918	0.9995	0.0497	1.006	0.0579	0.0179	35	-0.11	134.00	0
134	Serine (%)	0868	1.001	0.0020	1.006	0.0579	0.0179	35	-0.09	134.00	0
134	Serine (%)	0571	1.004	0.0160	1.006	0.0579	0.0179	35	-0.04	134.00	0
134	Serine (%)	0684	1.005	0.0100	1.006	0.0579	0.0179	35	-0.02	134.00	0
134	Serine (%)	0859	1.008	0.0265	1.006	0.0579	0.0179	35	0.04	134.00	0
134	Serine (%)	0910	1.010	0.0800	1.006	0.0579	0.0179	35	0.07	134.00	0
134	Serine (%)	0682	1.013	0.0000	1.006	0.0579	0.0179	35	0.12	134.00	0
134	Serine (%)	0354	1.016	0.0150	1.006	0.0579	0.0179	35	0.16	134.00	0
134	Serine (%)	0878	1.016	0.0180	1.006	0.0579	0.0179	35	0.17	134.00	0
134	Serine (%)	0872	1.020	0.0150	1.006	0.0579	0.0179	35	0.23	134.00	0
134	Serine (%)	2146	1.020	0.0200	1.006	0.0579	0.0179	35	0.24	134.99	0
134	Serine (%)	0227	1.040	0.0200	1.006	0.0579	0.0179	35	0.59	134.00	0
134	Serine (%)	0676	1.045	0.0700	1.006	0.0579	0.0179	35	0.67	134.05	0
134	Serine (%)	2196	1.046	0.0000	1.006	0.0579	0.0179	35	0.69	134.05	0
134	Serine (%)	0939	1.055	0.0100	1.006	0.0579	0.0179	35	0.84	134.00	0
134	Serine (%)	0626	1.072	0.0090	1.006	0.0579	0.0179	35	1.13	134.05	0
134	Serine (%)	0619	1.075	0.0300	1.006	0.0579	0.0179	35	1.19	134.00	0
134	Serine (%)	2144	1.085	0.0100	1.006	0.0579	0.0179	35	1.36	134.99	0
134	Serine (%)	0870	1.100	0.0020	1.006	0.0579	0.0179	35	1.62	134.00	0
134	Serine (%)	0171	1.129	0.0450	1.006	0.0579	0.0179	35	2.11	134.00	0
134	Serine (%)	0889	1.255	0.0100	1.006	0.0579	0.0179	35	4.30	134.99	0
134	Serine (%)	0226	1.152	0.1858	1.006	0.0579	0.0179	35	2.52	134.00	1
135	Threonine (%)	0407	0.6340	0.0220	0.8471	0.0334	0.0165	37	-6.38	135.05	0
135	Threonine (%)	0160	0.6350	0.0300	0.8471	0.0334	0.0165	37	-6.35	135.99	0
135	Threonine (%)	2144	0.7800	0.0400	0.8471	0.0334	0.0165	37	-2.01	135.99	0
135	Threonine (%)	0353	0.8000	0.0000	0.8471	0.0334	0.0165	37	-1.41	135.00	0
135	Threonine (%)	2188	0.8000	0.1200	0.8471	0.0334	0.0165	37	-1.41	135.05	0
135	Threonine (%)	0675	0.8100	0.0000	0.8471	0.0334	0.0165	37	-1.11	135.00	0
135	Threonine (%)	2256	0.8200	0.0020	0.8471	0.0334	0.0165	37	-0.81	135.00	0
135	Threonine (%)	0652	0.8200	0.0200	0.8471	0.0334	0.0165	37	-0.81	135.00	0
135	Threonine (%)	2059	0.8240	0.0060	0.8471	0.0334	0.0165	37	-0.69	135.00	0
135	Threonine (%)	0682	0.8260	0.0000	0.8471	0.0334	0.0165	37	-0.63	135.00	0
135	Threonine (%)	0968	0.8260	0.0040	0.8471	0.0334	0.0165	37	-0.63	135.00	0
135	Threonine (%)	2022	0.8300	0.0200	0.8471	0.0334	0.0165	37	-0.51	135.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
135	Threonine (%)	0859	0.8320	0.0235	0.8471	0.0334	0.0165	37	-0.45	135.00	0
135	Threonine (%)	0504	0.8400	0.0000	0.8471	0.0334	0.0165	37	-0.21	135.00	0
135	Threonine (%)	0918	0.8428	0.0268	0.8471	0.0334	0.0165	37	-0.13	135.00	0
135	Threonine (%)	0571	0.8435	0.0130	0.8471	0.0334	0.0165	37	-0.11	135.00	0
135	Threonine (%)	0644	0.8475	0.0050	0.8471	0.0334	0.0165	37	0.01	135.00	0
135	Threonine (%)	0868	0.8475	0.0070	0.8471	0.0334	0.0165	37	0.01	135.00	0
135	Threonine (%)	2246	0.8485	0.0549	0.8471	0.0334	0.0165	37	0.04	135.05	0
135	Threonine (%)	0619	0.8520	0.0200	0.8471	0.0334	0.0165	37	0.15	135.00	0
135	Threonine (%)	0354	0.8535	0.0110	0.8471	0.0334	0.0165	37	0.19	135.00	0
135	Threonine (%)	0878	0.8540	0.0200	0.8471	0.0334	0.0165	37	0.21	135.00	0
135	Threonine (%)	0610	0.8550	0.0100	0.8471	0.0334	0.0165	37	0.24	135.05	0
135	Threonine (%)	0098	0.8590	0.0060	0.8471	0.0334	0.0165	37	0.36	135.02	0
135	Threonine (%)	0676	0.8600	0.0400	0.8471	0.0334	0.0165	37	0.39	135.05	0
135	Threonine (%)	0872	0.8630	0.0140	0.8471	0.0334	0.0165	37	0.48	135.00	0
135	Threonine (%)	2146	0.8700	0.0000	0.8471	0.0334	0.0165	37	0.69	135.99	0
135	Threonine (%)	0227	0.8750	0.0100	0.8471	0.0334	0.0165	37	0.84	135.00	0
135	Threonine (%)	0910	0.8750	0.0100	0.8471	0.0334	0.0165	37	0.84	135.00	0
135	Threonine (%)	0939	0.8750	0.0300	0.8471	0.0334	0.0165	37	0.84	135.00	0
135	Threonine (%)	0626	0.8780	0.0040	0.8471	0.0334	0.0165	37	0.93	135.05	0
135	Threonine (%)	2115	0.8800	0.0200	0.8471	0.0334	0.0165	37	0.99	135.99	0
135	Threonine (%)	2196	0.8830	0.0000	0.8471	0.0334	0.0165	37	1.08	135.05	0
135	Threonine (%)	0684	0.8850	0.0100	0.8471	0.0334	0.0165	37	1.14	135.00	0
135	Threonine (%)	0171	0.8880	0.0100	0.8471	0.0334	0.0165	37	1.23	135.00	0
135	Threonine (%)	0870	0.8898	0.0005	0.8471	0.0334	0.0165	37	1.28	135.00	0
135	Threonine (%)	0889	0.9000	0.0000	0.8471	0.0334	0.0165	37	1.59	135.99	0
135	Threonine (%)	0226	1.039	0.3058	0.8471	0.0334	0.0165	37	5.74	135.00	1
136	Tryptophan (%)	0171	0.1360	0.0000	0.2562	0.0221	0.0079	22	-5.43	136.00	0
136	Tryptophan (%)	0504	0.2100	0.0400	0.2562	0.0221	0.0079	22	-2.09	136.99	0
136	Tryptophan (%)	0910	0.2200	0.0200	0.2562	0.0221	0.0079	22	-1.64	136.00	0
136	Tryptophan (%)	0878	0.2320	0.0060	0.2562	0.0221	0.0079	22	-1.09	136.01	0
136	Tryptophan (%)	0619	0.2460	0.0080	0.2562	0.0221	0.0079	22	-0.46	136.00	0
136	Tryptophan (%)	0868	0.2470	0.0080	0.2562	0.0221	0.0079	22	-0.42	136.01	0
136	Tryptophan (%)	0918	0.2500	0.0200	0.2562	0.0221	0.0079	22	-0.28	136.00	0
136	Tryptophan (%)	0407	0.2520	0.0240	0.2562	0.0221	0.0079	22	-0.19	136.05	0
136	Tryptophan (%)	0098	0.2525	0.0010	0.2562	0.0221	0.0079	22	-0.17	136.02	0
136	Tryptophan (%)	2114	0.2548	0.0049	0.2562	0.0221	0.0079	22	-0.07	136.01	0
136	Tryptophan (%)	0872	0.2555	0.0010	0.2562	0.0221	0.0079	22	-0.03	136.03	0
136	Tryptophan (%)	2059	0.2570	0.0040	0.2562	0.0221	0.0079	22	0.04	136.03	0
136	Tryptophan (%)	0644	0.2595	0.0010	0.2562	0.0221	0.0079	22	0.15	136.01	0
136	Tryptophan (%)	0859	0.2598	0.0022	0.2562	0.0221	0.0079	22	0.16	136.03	0
136	Tryptophan (%)	0571	0.2610	0.0020	0.2562	0.0221	0.0079	22	0.22	136.03	0
136	Tryptophan (%)	0870	0.2619	0.0026	0.2562	0.0221	0.0079	22	0.26	136.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
136	Tryptophan (%)	0939	0.2650	0.0100	0.2562	0.0221	0.0079	22	0.40	136.99	0
136	Tryptophan (%)	0353	0.2750	0.0100	0.2562	0.0221	0.0079	22	0.85	136.03	0
136	Tryptophan (%)	0227	0.2800	0.0000	0.2562	0.0221	0.0079	22	1.07	136.00	0
136	Tryptophan (%)	0684	0.2800	0.0000	0.2562	0.0221	0.0079	22	1.07	136.00	0
136	Tryptophan (%)	0682	0.4600	0.0000	0.2562	0.0221	0.0079	22	9.21	136.00	0
136	Tryptophan (%)	0889	0.5050	0.0100	0.2562	0.0221	0.0079	22	11.24	136.99	0
136	Tryptophan (%)	2246	0.4482	0.0713	0.2562	0.0221	0.0079	22	8.67	136.05	1
137	Tyrosine (%)	0682	0.5050	0.0000	0.6792	0.1187	0.0188	29	-1.47	137.00	0
137	Tyrosine (%)	0889	0.5050	0.0100	0.6792	0.1187	0.0188	29	-1.47	137.99	0
137	Tyrosine (%)	2144	0.5100	0.0000	0.6792	0.1187	0.0188	29	-1.43	137.99	0
137	Tyrosine (%)	0626	0.5270	0.0660	0.6792	0.1187	0.0188	29	-1.28	137.05	0
137	Tyrosine (%)	0968	0.5495	0.0630	0.6792	0.1187	0.0188	29	-1.09	137.00	0
137	Tyrosine (%)	0675	0.5650	0.0100	0.6792	0.1187	0.0188	29	-0.96	137.00	0
137	Tyrosine (%)	0910	0.6100	0.0600	0.6792	0.1187	0.0188	29	-0.58	137.00	0
137	Tyrosine (%)	0918	0.6150	0.0265	0.6792	0.1187	0.0188	29	-0.54	137.00	0
137	Tyrosine (%)	2256	0.6170	0.0160	0.6792	0.1187	0.0188	29	-0.52	137.00	0
137	Tyrosine (%)	0870	0.6350	0.0273	0.6792	0.1187	0.0188	29	-0.37	137.00	0
137	Tyrosine (%)	0160	0.6500	0.0200	0.6792	0.1187	0.0188	29	-0.25	137.99	0
137	Tyrosine (%)	0227	0.6600	0.0400	0.6792	0.1187	0.0188	29	-0.16	137.00	0
137	Tyrosine (%)	0684	0.6600	0.0000	0.6792	0.1187	0.0188	29	-0.16	137.00	0
137	Tyrosine (%)	0354	0.6735	0.0010	0.6792	0.1187	0.0188	29	-0.05	137.00	0
137	Tyrosine (%)	0226	0.6795	0.0114	0.6792	0.1187	0.0188	29	0.00	137.00	0
137	Tyrosine (%)	0644	0.6975	0.0110	0.6792	0.1187	0.0188	29	0.15	137.00	0
137	Tyrosine (%)	0872	0.7000	0.0120	0.6792	0.1187	0.0188	29	0.18	137.00	0
137	Tyrosine (%)	2196	0.7000	0.0000	0.6792	0.1187	0.0188	29	0.18	137.05	0
137	Tyrosine (%)	0171	0.7115	0.0150	0.6792	0.1187	0.0188	29	0.27	137.00	0
137	Tyrosine (%)	0504	0.7350	0.0100	0.6792	0.1187	0.0188	29	0.47	137.00	0
137	Tyrosine (%)	2059	0.7500	0.0040	0.6792	0.1187	0.0188	29	0.60	137.00	0
137	Tyrosine (%)	0098	0.7560	0.0000	0.6792	0.1187	0.0188	29	0.65	137.02	0
137	Tyrosine (%)	2188	0.7700	0.0200	0.6792	0.1187	0.0188	29	0.77	137.05	0
137	Tyrosine (%)	0353	0.7800	0.0000	0.6792	0.1187	0.0188	29	0.85	137.00	0
137	Tyrosine (%)	0676	0.7900	0.0200	0.6792	0.1187	0.0188	29	0.93	137.05	0
137	Tyrosine (%)	2146	0.7950	0.0300	0.6792	0.1187	0.0188	29	0.98	137.99	0
137	Tyrosine (%)	0610	0.8350	0.0100	0.6792	0.1187	0.0188	29	1.31	137.05	0
137	Tyrosine (%)	0407	0.8620	0.0320	0.6792	0.1187	0.0188	29	1.54	137.05	0
137	Tyrosine (%)	0939	0.9050	0.0300	0.6792	0.1187	0.0188	29	1.90	137.00	0
137	Tyrosine (%)	2246	0.5979	0.2122	0.6792	0.1187	0.0188	29	-0.68	137.05	1
138	Valine (%)	0226	0.8305	0.0816	1.083	0.0831	0.0260	38	-3.04	138.00	0
138	Valine (%)	0353	0.8950	0.0100	1.083	0.0831	0.0260	38	-2.26	138.00	0
138	Valine (%)	0889	0.8950	0.0100	1.083	0.0831	0.0260	38	-2.26	138.99	0
138	Valine (%)	2256	0.9535	0.0030	1.083	0.0831	0.0260	38	-1.56	138.00	0
138	Valine (%)	0675	0.9550	0.0100	1.083	0.0831	0.0260	38	-1.54	138.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
138	Valine (%)	0407	0.9580	0.0820	1.083	0.0831	0.0260	38	-1.50	138.05	0
138	Valine (%)	0859	1.014	0.0185	1.083	0.0831	0.0260	38	-0.83	138.00	0
138	Valine (%)	0619	1.030	0.0400	1.083	0.0831	0.0260	38	-0.64	138.00	0
138	Valine (%)	2188	1.040	0.0200	1.083	0.0831	0.0260	38	-0.52	138.05	0
138	Valine (%)	0160	1.040	0.0800	1.083	0.0831	0.0260	38	-0.52	138.99	0
138	Valine (%)	2144	1.040	0.0200	1.083	0.0831	0.0260	38	-0.52	138.99	0
138	Valine (%)	0868	1.054	0.0300	1.083	0.0831	0.0260	38	-0.35	138.00	0
138	Valine (%)	0939	1.055	0.0300	1.083	0.0831	0.0260	38	-0.34	138.00	0
138	Valine (%)	0354	1.057	0.0040	1.083	0.0831	0.0260	38	-0.31	138.00	0
138	Valine (%)	0968	1.059	0.0180	1.083	0.0831	0.0260	38	-0.29	138.00	0
138	Valine (%)	0652	1.065	0.0100	1.083	0.0831	0.0260	38	-0.22	138.00	0
138	Valine (%)	0872	1.067	0.0100	1.083	0.0831	0.0260	38	-0.19	138.00	0
138	Valine (%)	0870	1.078	0.0003	1.083	0.0831	0.0260	38	-0.05	138.00	0
138	Valine (%)	2196	1.085	0.0000	1.083	0.0831	0.0260	38	0.03	138.05	0
138	Valine (%)	0684	1.095	0.0100	1.083	0.0831	0.0260	38	0.15	138.00	0
138	Valine (%)	0910	1.100	0.0400	1.083	0.0831	0.0260	38	0.21	138.00	0
138	Valine (%)	0918	1.102	0.0604	1.083	0.0831	0.0260	38	0.23	138.00	0
138	Valine (%)	0227	1.105	0.0100	1.083	0.0831	0.0260	38	0.27	138.00	0
138	Valine (%)	2146	1.105	0.0100	1.083	0.0831	0.0260	38	0.27	138.99	0
138	Valine (%)	0571	1.112	0.0090	1.083	0.0831	0.0260	38	0.34	138.00	0
138	Valine (%)	2059	1.116	0.0050	1.083	0.0831	0.0260	38	0.39	138.00	0
138	Valine (%)	0171	1.140	0.0340	1.083	0.0831	0.0260	38	0.69	138.00	0
138	Valine (%)	0682	1.141	0.0000	1.083	0.0831	0.0260	38	0.70	138.00	0
138	Valine (%)	0676	1.145	0.0300	1.083	0.0831	0.0260	38	0.75	138.05	0
138	Valine (%)	0644	1.153	0.0010	1.083	0.0831	0.0260	38	0.84	138.00	0
138	Valine (%)	0098	1.158	0.0040	1.083	0.0831	0.0260	38	0.90	138.02	0
138	Valine (%)	0504	1.165	0.0100	1.083	0.0831	0.0260	38	0.99	138.00	0
138	Valine (%)	0878	1.166	0.1130	1.083	0.0831	0.0260	38	0.99	138.00	0
138	Valine (%)	0626	1.174	0.0200	1.083	0.0831	0.0260	38	1.10	138.05	0
138	Valine (%)	2246	1.179	0.0628	1.083	0.0831	0.0260	38	1.16	138.05	0
138	Valine (%)	2022	1.180	0.0200	1.083	0.0831	0.0260	38	1.17	138.00	0
138	Valine (%)	2115	1.183	0.0510	1.083	0.0831	0.0260	38	1.20	138.99	0
138	Valine (%)	0610	1.200	0.0200	1.083	0.0831	0.0260	38	1.41	138.05	0
139	Taurine (%)	0171	0.0295	0.0010	0.0621	0.0493	0.0010	7	-0.66	139.00	0
139	Taurine (%)	0227	0.0300	0.0000	0.0621	0.0493	0.0010	7	-0.65	139.00	0
139	Taurine (%)	0098	0.0350	0.0040	0.0621	0.0493	0.0010	7	-0.55	139.02	0
139	Taurine (%)	0407	0.0420	0.0020	0.0621	0.0493	0.0010	7	-0.41	139.05	0
139	Taurine (%)	0889	0.0500	0.0000	0.0621	0.0493	0.0010	7	-0.24	139.99	0
139	Taurine (%)	0682	0.1120	0.0000	0.0621	0.0493	0.0010	7	1.01	139.00	0
139	Taurine (%)	0504	0.2200	0.0000	0.0621	0.0493	0.0010	7	3.21	139.00	0
139	Taurine (%)	0676	22.33	0.7700	0.0621	0.0493	0.0010	7	451.91	139.05	2
160	Fructose (%)	0619	0.0725	0.0010				2		160.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
160	Fructose (%)	0227	0.1950	0.0500				2		160.99	0
162	Glucose (%)	0619	0.0630	0.0000				1		162.99	0
162	Glucose (%)	0227	< 0.15					1		162.99	5
163	Lactose (%)	0619	0.0000	0.0000				0		163.99	4
163	Lactose (%)	0227	< 0.15					0		163.99	5
164	Maltose (%)	0227	< 0.15					0		164.99	5
165	Sucrose (%)	0619	1.710	0.0200				2		165.99	0
165	Sucrose (%)	0227	1.730	0.0400				2		165.99	0
166	Raffinose (%)	0619	0.4395	0.0030				2		166.99	0
166	Raffinose (%)	0227	0.4400	0.0200				2		166.99	0
167	Stachyose (%)	0227	0.4950	0.0500				2		167.99	0
167	Stachyose (%)	0619	0.6410	0.0000				2		167.99	0
400	Water Activity (Units)	0942	0.3885	0.0010	0.4146	0.0255	0.0025	10	-1.03	400.01	0
400	Water Activity (Units)	2073	0.3975	0.0030	0.4146	0.0255	0.0025	10	-0.67	400.99	0
400	Water Activity (Units)	0843	0.3985	0.0010	0.4146	0.0255	0.0025	10	-0.63	400.99	0
400	Water Activity (Units)	0843	0.4000	0.0055	0.4146	0.0255	0.0025	10	-0.58	400.01	0
400	Water Activity (Units)	0051	0.4070	0.0020	0.4146	0.0255	0.0025	10	-0.30	400.01	0
400	Water Activity (Units)	0407	0.4091	0.0007	0.4146	0.0255	0.0025	10	-0.22	400.01	0
400	Water Activity (Units)	0589	0.4115	0.0030	0.4146	0.0255	0.0025	10	-0.12	400.01	0
400	Water Activity (Units)	2181	0.4313	0.0025	0.4146	0.0255	0.0025	10	0.65	400.01	0
400	Water Activity (Units)	2083	0.4500	0.0000	0.4146	0.0255	0.0025	10	1.39	400.99	0
400	Water Activity (Units)	2109	0.5859	0.0059	0.4146	0.0255	0.0025	10	6.73	400.01	0
412	Starch, Dietary (%)	0353	31.17	0.1200				1		412.01	0
516	Arsenic, Total (mg / kg (ppm))	0045	0.0376	0.0079	0.0680	0.0283	0.0043	6	-1.08	516.00	0
516	Arsenic, Total (mg / kg (ppm))	0910	0.0550	0.0100	0.0680	0.0283	0.0043	6	-0.46	516.52	0
516	Arsenic, Total (mg / kg (ppm))	0171	0.0575	0.0010	0.0680	0.0283	0.0043	6	-0.37	516.00	0
516	Arsenic, Total (mg / kg (ppm))	0208	0.0696	0.0060	0.0680	0.0283	0.0043	6	0.06	516.52	0
516	Arsenic, Total (mg / kg (ppm))	2012	0.0780	0.0000	0.0680	0.0283	0.0043	6	0.35	516.53	0
516	Arsenic, Total (mg / kg (ppm))	0918	0.1378	0.0010	0.0680	0.0283	0.0043	6	2.47	516.53	0
516	Arsenic, Total (mg / kg (ppm))	0407	1.078	0.2245	0.0680	0.0283	0.0043	6	35.72	516.42	1
516	Arsenic, Total (mg / kg (ppm))	2124	< 0.89		0.0680	0.0283	0.0043	6		516.43	5
518	Cadmium (mg / kg (ppm))	2246	0.0495	0.0022	0.0640	0.0073	0.0041	8	-2.00	518.34	0
518	Cadmium (mg / kg (ppm))	0918	0.0607	0.0029	0.0640	0.0073	0.0041	8	-0.46	518.53	0
518	Cadmium (mg / kg (ppm))	0208	0.0609	0.0044	0.0640	0.0073	0.0041	8	-0.43	518.52	0
518	Cadmium (mg / kg (ppm))	0186	0.0610	0.0005	0.0640	0.0073	0.0041	8	-0.42	518.52	0
518	Cadmium (mg / kg (ppm))	0910	0.0650	0.0100	0.0640	0.0073	0.0041	8	0.14	518.52	0
518	Cadmium (mg / kg (ppm))	0968	0.0680	0.0020	0.0640	0.0073	0.0041	8	0.55	518.43	0
518	Cadmium (mg / kg (ppm))	0407	0.0703	0.0048	0.0640	0.0073	0.0041	8	0.87	518.41	0
518	Cadmium (mg / kg (ppm))	2012	0.0730	0.0060	0.0640	0.0073	0.0041	8	1.24	518.53	0
518	Cadmium (mg / kg (ppm))	0171	< 0.05		0.0640	0.0073	0.0041	8		518.41	5
518	Cadmium (mg / kg (ppm))	2124	< 0.16		0.0640	0.0073	0.0041	8		518.43	5
518	Cadmium (mg / kg (ppm))	0689	< 0.56		0.0640	0.0073	0.0041	8		518.31	5

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
520	Chromium (mg / kg (ppm))	0186	0.7900	0.0560	1.964	0.8419	0.2448	9	-1.39	520.52	0
520	Chromium (mg / kg (ppm))	0968	1.054	0.0300	1.964	0.8419	0.2448	9	-1.08	520.43	0
520	Chromium (mg / kg (ppm))	0407	1.557	0.0267	1.964	0.8419	0.2448	9	-0.48	520.41	0
520	Chromium (mg / kg (ppm))	0918	1.611	0.3568	1.964	0.8419	0.2448	9	-0.42	520.53	0
520	Chromium (mg / kg (ppm))	0171	2.215	0.0700	1.964	0.8419	0.2448	9	0.30	520.41	0
520	Chromium (mg / kg (ppm))	0164	2.325	0.2100	1.964	0.8419	0.2448	9	0.43	520.53	0
520	Chromium (mg / kg (ppm))	0693	2.585	0.5540	1.964	0.8419	0.2448	9	0.74	520.42	0
520	Chromium (mg / kg (ppm))	0045	2.640	0.7000	1.964	0.8419	0.2448	9	0.80	520.42	0
520	Chromium (mg / kg (ppm))	0510	2.900	0.2000	1.964	0.8419	0.2448	9	1.11	520.43	0
526	Lead (mg / kg (ppm))	0171	1.145	0.0100	1.357	0.1935	0.0862	10	-1.10	526.41	0
526	Lead (mg / kg (ppm))	0407	1.156	0.0713	1.357	0.1935	0.0862	10	-1.04	526.41	0
526	Lead (mg / kg (ppm))	0968	1.191	0.0370	1.357	0.1935	0.0862	10	-0.86	526.43	0
526	Lead (mg / kg (ppm))	0186	1.325	0.1100	1.357	0.1935	0.0862	10	-0.17	526.52	0
526	Lead (mg / kg (ppm))	2055	1.350	0.1000	1.357	0.1935	0.0862	10	-0.04	526.43	0
526	Lead (mg / kg (ppm))	0208	1.350	0.0400	1.357	0.1935	0.0862	10	-0.04	526.52	0
526	Lead (mg / kg (ppm))	0910	1.365	0.0500	1.357	0.1935	0.0862	10	0.04	526.52	0
526	Lead (mg / kg (ppm))	2012	1.450	0.0970	1.357	0.1935	0.0862	10	0.48	526.53	0
526	Lead (mg / kg (ppm))	2246	1.592	0.2670	1.357	0.1935	0.0862	10	1.21	526.34	0
526	Lead (mg / kg (ppm))	0918	1.688	0.0798	1.357	0.1935	0.0862	10	1.71	526.53	0
526	Lead (mg / kg (ppm))	0689	< 0.2		1.357	0.1935	0.0862	10		526.31	5
529	Mercury (µg / kg (ppb))	2012	5.980	0.0000				1		529.99	0
539	Nickel (mg / kg (ppm))	0186	1.775	0.2300	2.000	0.2209	0.0863	4		539.52	0
539	Nickel (mg / kg (ppm))	0407	1.908	0.0716	2.000	0.2209	0.0863	4		539.41	0
539	Nickel (mg / kg (ppm))	0171	2.020	0.0200	2.000	0.2209	0.0863	4		539.41	0
539	Nickel (mg / kg (ppm))	0918	2.295	0.0237	2.000	0.2209	0.0863	4		539.53	0
710	Lauric Acid (12:0) (% (w/w))	0226	0.0055	0.0003				2		710.99	0
710	Lauric Acid (12:0) (% (w/w))	0676	0.0055	0.0010				2		710.99	0
710	Lauric Acid (12:0) (% (w/w))	2124	< 0.01					2		710.99	5
710	Lauric Acid (12:0) (% (w/w))	0164	< 0.02					2		710.99	5
714	Myristic Acid (14:0) (% (w/w))	2124	0.0550	0.0100				3		714.99	0
714	Myristic Acid (14:0) (% (w/w))	0226	0.0741	0.0025				3		714.99	0
714	Myristic Acid (14:0) (% (w/w))	0164	0.0800	0.0000				3		714.99	0
714	Myristic Acid (14:0) (% (w/w))	0910	0.8600	0.0200				3		714.99	2
716	Palmitic Acid (16:0) (% (w/w))	2124	1.340	0.0600				3		716.99	0
716	Palmitic Acid (16:0) (% (w/w))	0164	1.735	0.0300				3		716.99	0
716	Palmitic Acid (16:0) (% (w/w))	0226	1.762	0.0525				3		716.99	0
716	Palmitic Acid (16:0) (% (w/w))	0910	20.73	0.0800				3		716.99	2
718	Palmitoleic Acid (9c-16:1) (% (w/w))	2124	0.2150	0.0100	0.2538	0.0307	0.0047	4		718.99	0
718	Palmitoleic Acid (9c-16:1) (% (w/w))	0676	0.2525	0.0010	0.2538	0.0307	0.0047	4		718.99	0
718	Palmitoleic Acid (9c-16:1) (% (w/w))	0226	0.2578	0.0077	0.2538	0.0307	0.0047	4		718.99	0
718	Palmitoleic Acid (9c-16:1) (% (w/w))	0164	0.2900	0.0000	0.2538	0.0307	0.0047	4		718.99	0
718	Palmitoleic Acid (9c-16:1) (% (w/w))	0910	3.460	0.0200	0.2538	0.0307	0.0047	4		718.99	2

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
722	Stearic Acid (18:0) (% (w/w))	2124	0.4500	0.0200				3	722.99	0	
722	Stearic Acid (18:0) (% (w/w))	0164	0.5750	0.0100				3	722.99	0	
722	Stearic Acid (18:0) (% (w/w))	0226	0.6094	0.0198				3	722.99	0	
722	Stearic Acid (18:0) (% (w/w))	0910	6.970	0.0600				3	722.99	2	
724	Oleic Acid (9c-18:1) (% (w/w))	2124	2.125	0.0900				3	724.99	0	
724	Oleic Acid (9c-18:1) (% (w/w))	0164	2.460	0.0400				3	724.99	0	
724	Oleic Acid (9c-18:1) (% (w/w))	0226	2.613	0.0206				3	724.99	0	
724	Oleic Acid (9c-18:1) (% (w/w))	0910	34.10	0.2300				3	724.99	2	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	2124	1.820	0.0800	2.201	0.3009	0.0491	4	726.99	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0164	2.165	0.0500	2.201	0.3009	0.0491	4	726.99	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0226	2.270	0.0564	2.201	0.3009	0.0491	4	726.99	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0676	2.548	0.0100	2.201	0.3009	0.0491	4	726.99	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0910	27.88	0.0500	2.201	0.3009	0.0491	4	726.99	2	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/w))	2124	0.1000	0.0000	0.1256	0.0210	0.0041	4	728.99	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/w))	0164	0.1250	0.0100	0.1256	0.0210	0.0041	4	728.99	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/w))	0226	0.1260	0.0033	0.1256	0.0210	0.0041	4	728.99	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/w))	0676	0.1515	0.0030	0.1256	0.0210	0.0041	4	728.99	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/w))	0910	1.620	0.0200	0.1256	0.0210	0.0041	4	728.99	2	
730	Arachidic Acid (20:0) (% (w/w))	0226	0.0158	0.0006				2	730.99	0	
730	Arachidic Acid (20:0) (% (w/w))	0910	0.2050	0.0100				2	730.99	0	
730	Arachidic Acid (20:0) (% (w/w))	0164	< 0.02					2	730.99	5	
732	Gondoic Acid (11c-20:1) (% (w/w))	0164	0.0300	0.0000				3	732.99	0	
732	Gondoic Acid (11c-20:1) (% (w/w))	0226	0.0359	0.0008				3	732.99	0	
732	Gondoic Acid (11c-20:1) (% (w/w))	0910	0.5450	0.0100				3	732.99	0	
736	Arachidonic Acid (5c,8c,11c,14c-20:4) (% (w/w))	0676	0.0405	0.0030				2	736.99	0	
736	Arachidonic Acid (5c,8c,11c,14c-20:4) (% (w/w))	0910	0.2650	0.0300				2	736.99	0	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c) (% (w/w))	0226	0.0070	0.0006				3	740.99	0	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c) (% (w/w))	0676	0.0070	0.0020				3	740.99	0	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c) (% (w/w))	2124	0.0100	0.0000				3	740.99	0	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c) (% (w/w))	0910	0.0950	0.0100				3	740.99	2	
742	Behenic Acid (22:0) (% (w/w))	0226	0.0078	0.0004				2	742.99	0	
742	Behenic Acid (22:0) (% (w/w))	0910	0.1050	0.0100				2	742.99	0	
742	Behenic Acid (22:0) (% (w/w))	0164	< 0.02					2	742.99	5	
744	Erucic Acid (13c-22:1) (% (w/w))	0226	0.0000	0.0000				0	744.99	4	
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c) (% (w/w))	0226	0.0046	0.0001				2	746.99	0	
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c) (% (w/w))	0676	0.0060	0.0000				2	746.99	0	
748	Lignoceric Acid (24:0) (% (w/w))	0226	0.0092	0.0007				2	748.99	0	
748	Lignoceric Acid (24:0) (% (w/w))	0910	0.1000	0.0600				2	748.99	0	
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22c) (% (w/w))	0226	0.0057	0.0001				2	750.99	0	
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22c) (% (w/w))	0676	0.0070	0.0000				2	750.99	0	
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22c) (% (w/w))	2124	< 0.01					2	750.99	5	
752	Nervonic Acid (24:1) isomers (% (w/w))	0226	0.0030	0.0002				1	752.99	0	

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
754	Total n-3 Polyunsaturated (Omega-3) Fatty Aci	2124	0.1200	0.0000				3	754.99	0	
754	Total n-3 Polyunsaturated (Omega-3) Fatty Aci	0164	0.1450	0.0100				3	754.99	0	
754	Total n-3 Polyunsaturated (Omega-3) Fatty Aci	0676	0.1700	0.0000				3	754.99	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Aci	2124	1.880	0.0800				3	756.99	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Aci	0164	2.205	0.0700				3	756.99	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Aci	0676	2.625	0.0100				3	756.99	0	
758	Total Saturated Fatty Acids (% (w/w))	0164	2.485	0.0300				2	758.99	0	
758	Total Saturated Fatty Acids (% (w/w))	0910	29.42	0.0500				2	758.99	0	
762	Total Monounsaturated Fatty Acids (% (w/w))	0164	2.975	0.0300				2	762.99	0	
762	Total Monounsaturated Fatty Acids (% (w/w))	0910	38.47	0.2200				2	762.99	0	
766	Total Polyunsaturated Fatty Acids (% (w/w))	0164	2.385	0.0700				2	766.99	0	
766	Total Polyunsaturated Fatty Acids (% (w/w))	0910	29.92	0.0600				2	766.99	0	
770	Total Fat (equivalent to NLEA) (% (w/w))	0164	8.300	0.1400				1	770.99	0	
772	Total Fatty Acids (% (w/w))	2124	6.230	0.2800				3	772.99	0	
772	Total Fatty Acids (% (w/w))	0164	7.935	0.1300				3	772.99	0	
772	Total Fatty Acids (% (w/w))	0226	8.019	0.1993				3	772.99	0	

Notes: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2 and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = Rejected for duplicates too far apart, 2 = Rejected as outlier, 5 = A reporting limit and 4 = zeros submitted as values. Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). Flag 3 indicates not used in statistics.