

Animal Feed Scheme
Goat Feed, Medicated

Labs Reporting: 200
Analytes Reported 109
Issue Date : 05/31/2019

Test Material Code # 201924

Analyte Proficiency Testing Report

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
000	Urea (%)	0278	1.000	0.0000				1		000.02	0
000	Urea (%)	2246	< 0.32					1		000.02	5
001	Loss on Drying (%)	0618	6.315	0.1500	8.243	0.2828	0.1024	87	-6.82	001.99	0
001	Loss on Drying (%)	0505	6.935	0.2900	8.243	0.2828	0.1024	87	-4.62	001.07	0
001	Loss on Drying (%)	0297	7.170	0.0000	8.243	0.2828	0.1024	87	-3.79	001.07	0
001	Loss on Drying (%)	2259	7.367	0.0830	8.243	0.2828	0.1024	87	-3.10	001.07	0
001	Loss on Drying (%)	0074	7.445	0.0300	8.243	0.2828	0.1024	87	-2.82	001.07	0
001	Loss on Drying (%)	0345	7.495	0.0100	8.243	0.2828	0.1024	87	-2.64	001.07	0
001	Loss on Drying (%)	0918	7.805	0.0500	8.243	0.2828	0.1024	87	-1.55	001.99	0
001	Loss on Drying (%)	0083	7.825	0.2300	8.243	0.2828	0.1024	87	-1.48	001.07	0
001	Loss on Drying (%)	0038	7.885	0.0100	8.243	0.2828	0.1024	87	-1.26	001.07	0
001	Loss on Drying (%)	2246	7.890	0.0600	8.243	0.2828	0.1024	87	-1.25	001.99	0
001	Loss on Drying (%)	0034	7.905	0.1500	8.243	0.2828	0.1024	87	-1.19	001.07	0
001	Loss on Drying (%)	2146	7.915	0.0100	8.243	0.2828	0.1024	87	-1.16	001.07	0
001	Loss on Drying (%)	0644	7.960	0.2000	8.243	0.2828	0.1024	87	-1.00	001.05	0
001	Loss on Drying (%)	2222	7.971	0.0220	8.243	0.2828	0.1024	87	-0.96	001.07	0
001	Loss on Drying (%)	2155	7.975	0.3700	8.243	0.2828	0.1024	87	-0.95	001.99	0
001	Loss on Drying (%)	0676	7.995	0.0500	8.243	0.2828	0.1024	87	-0.88	001.99	0
001	Loss on Drying (%)	0675	8.000	0.0600	8.243	0.2828	0.1024	87	-0.86	001.07	0
001	Loss on Drying (%)	0186	8.000	0.0200	8.243	0.2828	0.1024	87	-0.86	001.99	0
001	Loss on Drying (%)	2081	8.000	0.0600	8.243	0.2828	0.1024	87	-0.86	001.99	0
001	Loss on Drying (%)	0610	8.030	0.0200	8.243	0.2828	0.1024	87	-0.75	001.05	0
001	Loss on Drying (%)	0015	8.040	0.2600	8.243	0.2828	0.1024	87	-0.72	001.07	0
001	Loss on Drying (%)	2190	8.050	0.0200	8.243	0.2828	0.1024	87	-0.68	001.07	0
001	Loss on Drying (%)	0027	8.051	0.0170	8.243	0.2828	0.1024	87	-0.68	001.00	0
001	Loss on Drying (%)	0536	8.070	0.0200	8.243	0.2828	0.1024	87	-0.61	001.99	0
001	Loss on Drying (%)	0187	8.085	0.0100	8.243	0.2828	0.1024	87	-0.56	001.07	0
001	Loss on Drying (%)	0853	8.100	0.2000	8.243	0.2828	0.1024	87	-0.50	001.07	0
001	Loss on Drying (%)	0560	8.100	0.0600	8.243	0.2828	0.1024	87	-0.50	001.99	0
001	Loss on Drying (%)	0571	8.125	0.0500	8.243	0.2828	0.1024	87	-0.42	001.07	0
001	Loss on Drying (%)	0849	8.127	0.0740	8.243	0.2828	0.1024	87	-0.41	001.07	0
001	Loss on Drying (%)	0590	8.130	0.0600	8.243	0.2828	0.1024	87	-0.40	001.07	0
001	Loss on Drying (%)	0683	8.130	0.0400	8.243	0.2828	0.1024	87	-0.40	001.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
001	Loss on Drying (%)	0948	8.145	0.0100	8.243	0.2828	0.1024	87	-0.35	001.99	0
001	Loss on Drying (%)	0169	8.165	0.0300	8.243	0.2828	0.1024	87	-0.27	001.00	0
001	Loss on Drying (%)	0682	8.180	0.0000	8.243	0.2828	0.1024	87	-0.22	001.07	0
001	Loss on Drying (%)	0028	8.185	0.0100	8.243	0.2828	0.1024	87	-0.20	001.07	0
001	Loss on Drying (%)	0142	8.195	0.0100	8.243	0.2828	0.1024	87	-0.17	001.07	0
001	Loss on Drying (%)	0089	8.200	0.0200	8.243	0.2828	0.1024	87	-0.15	001.07	0
001	Loss on Drying (%)	0510	8.200	0.2000	8.243	0.2828	0.1024	87	-0.15	001.99	0
001	Loss on Drying (%)	0910	8.200	0.2000	8.243	0.2828	0.1024	87	-0.15	001.99	0
001	Loss on Drying (%)	0045	8.205	0.3300	8.243	0.2828	0.1024	87	-0.13	001.07	0
001	Loss on Drying (%)	0720	8.220	0.0400	8.243	0.2828	0.1024	87	-0.08	001.99	0
001	Loss on Drying (%)	2059	8.225	0.0500	8.243	0.2828	0.1024	87	-0.06	001.07	0
001	Loss on Drying (%)	0889	8.225	0.2500	8.243	0.2828	0.1024	87	-0.06	001.99	0
001	Loss on Drying (%)	2018	8.230	0.0200	8.243	0.2828	0.1024	87	-0.04	001.99	0
001	Loss on Drying (%)	0226	8.240	0.0600	8.243	0.2828	0.1024	87	-0.01	001.07	0
001	Loss on Drying (%)	0035	8.245	0.1900	8.243	0.2828	0.1024	87	0.01	001.07	0
001	Loss on Drying (%)	0037	8.250	0.0600	8.243	0.2828	0.1024	87	0.03	001.99	0
001	Loss on Drying (%)	2264	8.263	0.1506	8.243	0.2828	0.1024	87	0.07	001.07	0
001	Loss on Drying (%)	0868	8.275	0.0100	8.243	0.2828	0.1024	87	0.11	001.03	0
001	Loss on Drying (%)	0612	8.280	0.5200	8.243	0.2828	0.1024	87	0.13	001.99	0
001	Loss on Drying (%)	2126	8.290	0.0200	8.243	0.2828	0.1024	87	0.17	001.99	0
001	Loss on Drying (%)	0852	8.315	0.1300	8.243	0.2828	0.1024	87	0.26	001.99	0
001	Loss on Drying (%)	0278	8.330	0.0200	8.243	0.2828	0.1024	87	0.31	001.07	0
001	Loss on Drying (%)	0049	8.335	0.0100	8.243	0.2828	0.1024	87	0.33	001.07	0
001	Loss on Drying (%)	2109	8.350	0.0400	8.243	0.2828	0.1024	87	0.38	001.07	0
001	Loss on Drying (%)	0619	8.355	0.0100	8.243	0.2828	0.1024	87	0.40	001.03	0
001	Loss on Drying (%)	0504	8.360	0.3000	8.243	0.2828	0.1024	87	0.41	001.00	0
001	Loss on Drying (%)	0695	8.370	0.1600	8.243	0.2828	0.1024	87	0.45	001.07	0
001	Loss on Drying (%)	0693	8.382	0.0250	8.243	0.2828	0.1024	87	0.49	001.07	0
001	Loss on Drying (%)	0309	8.391	0.1350	8.243	0.2828	0.1024	87	0.52	001.00	0
001	Loss on Drying (%)	0148	8.395	0.0100	8.243	0.2828	0.1024	87	0.54	001.07	0
001	Loss on Drying (%)	2242	8.400	0.4000	8.243	0.2828	0.1024	87	0.56	001.99	0
001	Loss on Drying (%)	0033	8.425	0.0100	8.243	0.2828	0.1024	87	0.64	001.07	0
001	Loss on Drying (%)	0139	8.430	0.0400	8.243	0.2828	0.1024	87	0.66	001.07	0
001	Loss on Drying (%)	0656	8.435	0.1500	8.243	0.2828	0.1024	87	0.68	001.99	0
001	Loss on Drying (%)	0689	8.450	0.1000	8.243	0.2828	0.1024	87	0.73	001.07	0
001	Loss on Drying (%)	2192	8.460	0.4000	8.243	0.2828	0.1024	87	0.77	001.07	0
001	Loss on Drying (%)	0878	8.465	0.1500	8.243	0.2828	0.1024	87	0.79	001.03	0
001	Loss on Drying (%)	0171	8.480	0.0000	8.243	0.2828	0.1024	87	0.84	001.07	0
001	Loss on Drying (%)	0876	8.480	0.2400	8.243	0.2828	0.1024	87	0.84	001.07	0
001	Loss on Drying (%)	2053	8.485	0.0700	8.243	0.2828	0.1024	87	0.86	001.07	0
001	Loss on Drying (%)	0638	8.500	0.0000	8.243	0.2828	0.1024	87	0.91	001.07	0
001	Loss on Drying (%)	0353	8.510	0.2000	8.243	0.2828	0.1024	87	0.95	001.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
001	Loss on Drying (%)	0893	8.515	0.1100	8.243	0.2828	0.1024	87	0.96	001.03	0
001	Loss on Drying (%)	2062	8.524	0.0173	8.243	0.2828	0.1024	87	1.00	001.03	0
001	Loss on Drying (%)	0066	8.540	0.1000	8.243	0.2828	0.1024	87	1.05	001.07	0
001	Loss on Drying (%)	0098	8.545	0.1300	8.243	0.2828	0.1024	87	1.07	001.07	0
001	Loss on Drying (%)	0003	8.555	0.0500	8.243	0.2828	0.1024	87	1.10	001.07	0
001	Loss on Drying (%)	0581	8.585	0.0100	8.243	0.2828	0.1024	87	1.21	001.07	0
001	Loss on Drying (%)	0872	8.605	0.0700	8.243	0.2828	0.1024	87	1.28	001.07	0
001	Loss on Drying (%)	0407	8.634	0.0236	8.243	0.2828	0.1024	87	1.38	001.07	0
001	Loss on Drying (%)	2258	8.730	0.1000	8.243	0.2828	0.1024	87	1.72	001.99	0
001	Loss on Drying (%)	0019	8.745	0.0300	8.243	0.2828	0.1024	87	1.78	001.99	0
001	Loss on Drying (%)	0843	8.830	0.3200	8.243	0.2828	0.1024	87	2.08	001.03	0
001	Loss on Drying (%)	0843	8.830	0.3200	8.243	0.2828	0.1024	87	2.08	001.07	0
001	Loss on Drying (%)	0357	9.010	0.0520	8.243	0.2828	0.1024	87	2.71	001.99	0
001	Loss on Drying (%)	0529	9.385	0.1100	8.243	0.2828	0.1024	87	4.04	001.07	0
001	Loss on Drying (%)	0366	9.000	1.000	8.243	0.2828	0.1024	87	2.68	001.07	1
002	Protein, Crude (%)	0047	15.88	0.1700	16.75	0.3238	0.1552	190	-2.70	002.06	0
002	Protein, Crude (%)	2128	15.92	0.3650	16.75	0.3238	0.1552	190	-2.54	002.03	0
002	Protein, Crude (%)	2259	16.07	0.2380	16.75	0.3238	0.1552	190	-2.10	002.01	0
002	Protein, Crude (%)	0529	16.11	0.3400	16.75	0.3238	0.1552	190	-1.97	002.06	0
002	Protein, Crude (%)	0169	16.13	0.4100	16.75	0.3238	0.1552	190	-1.92	002.02	0
002	Protein, Crude (%)	0139	16.14	0.0300	16.75	0.3238	0.1552	190	-1.89	002.06	0
002	Protein, Crude (%)	2188	16.15	0.0400	16.75	0.3238	0.1552	190	-1.85	002.01	0
002	Protein, Crude (%)	0870	16.18	0.3710	16.75	0.3238	0.1552	190	-1.75	002.01	0
002	Protein, Crude (%)	2146	16.19	0.1800	16.75	0.3238	0.1552	190	-1.72	002.05	0
002	Protein, Crude (%)	0644	16.21	0.2790	16.75	0.3238	0.1552	190	-1.68	002.06	0
002	Protein, Crude (%)	0598	16.22	0.2800	16.75	0.3238	0.1552	190	-1.63	002.06	0
002	Protein, Crude (%)	0539	16.23	0.1700	16.75	0.3238	0.1552	190	-1.62	002.06	0
002	Protein, Crude (%)	0910	16.25	0.5000	16.75	0.3238	0.1552	190	-1.54	002.01	0
002	Protein, Crude (%)	2190	16.26	0.3300	16.75	0.3238	0.1552	190	-1.52	002.05	0
002	Protein, Crude (%)	2155	16.26	0.0690	16.75	0.3238	0.1552	190	-1.52	002.06	0
002	Protein, Crude (%)	2246	16.27	0.1800	16.75	0.3238	0.1552	190	-1.48	002.05	0
002	Protein, Crude (%)	0852	16.29	0.1700	16.75	0.3238	0.1552	190	-1.43	002.05	0
002	Protein, Crude (%)	0011	16.33	0.2500	16.75	0.3238	0.1552	190	-1.31	002.06	0
002	Protein, Crude (%)	0874	16.34	0.2600	16.75	0.3238	0.1552	190	-1.26	002.04	0
002	Protein, Crude (%)	0675	16.34	0.0400	16.75	0.3238	0.1552	190	-1.26	002.05	0
002	Protein, Crude (%)	2199	16.35	0.3000	16.75	0.3238	0.1552	190	-1.23	002.05	0
002	Protein, Crude (%)	0294	16.35	0.1000	16.75	0.3238	0.1552	190	-1.23	002.06	0
002	Protein, Crude (%)	0720	16.36	0.0300	16.75	0.3238	0.1552	190	-1.21	002.06	0
002	Protein, Crude (%)	2264	16.37	0.1550	16.75	0.3238	0.1552	190	-1.17	002.05	0
002	Protein, Crude (%)	0650	16.38	0.0500	16.75	0.3238	0.1552	190	-1.15	002.06	0
002	Protein, Crude (%)	0345	16.38	0.0400	16.75	0.3238	0.1552	190	-1.14	002.06	0
002	Protein, Crude (%)	0885	16.38	0.0243	16.75	0.3238	0.1552	190	-1.13	002.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein, Crude (%)	2192	16.39	0.0300	16.75	0.3238	0.1552	190	-1.12	002.05	0
002	Protein, Crude (%)	2258	16.40	0.1700	16.75	0.3238	0.1552	190	-1.09	002.05	0
002	Protein, Crude (%)	0889	16.40	0.0800	16.75	0.3238	0.1552	190	-1.08	002.01	0
002	Protein, Crude (%)	0955	16.40	0.4000	16.75	0.3238	0.1552	190	-1.08	002.05	0
002	Protein, Crude (%)	0144	16.41	0.2200	16.75	0.3238	0.1552	190	-1.04	002.06	0
002	Protein, Crude (%)	2181	16.43	0.3300	16.75	0.3238	0.1552	190	-1.00	002.05	0
002	Protein, Crude (%)	0849	16.43	0.0320	16.75	0.3238	0.1552	190	-0.97	002.05	0
002	Protein, Crude (%)	0683	16.44	0.1400	16.75	0.3238	0.1552	190	-0.95	002.05	0
002	Protein, Crude (%)	0510	16.45	0.1000	16.75	0.3238	0.1552	190	-0.92	002.06	0
002	Protein, Crude (%)	0100	16.46	0.0300	16.75	0.3238	0.1552	190	-0.91	002.06	0
002	Protein, Crude (%)	2073	16.47	0.0100	16.75	0.3238	0.1552	190	-0.87	002.05	0
002	Protein, Crude (%)	0964	16.47	0.0910	16.75	0.3238	0.1552	190	-0.87	002.06	0
002	Protein, Crude (%)	0042	16.48	0.1500	16.75	0.3238	0.1552	190	-0.84	002.02	0
002	Protein, Crude (%)	0536	16.48	0.6100	16.75	0.3238	0.1552	190	-0.84	002.05	0
002	Protein, Crude (%)	2258	16.51	0.2200	16.75	0.3238	0.1552	190	-0.74	002.06	0
002	Protein, Crude (%)	0119	16.52	0.2860	16.75	0.3238	0.1552	190	-0.70	002.06	0
002	Protein, Crude (%)	0581	16.54	0.1500	16.75	0.3238	0.1552	190	-0.66	002.06	0
002	Protein, Crude (%)	0505	16.54	0.0200	16.75	0.3238	0.1552	190	-0.64	002.06	0
002	Protein, Crude (%)	2062	16.54	0.0150	16.75	0.3238	0.1552	190	-0.64	002.05	0
002	Protein, Crude (%)	2109	16.55	0.0600	16.75	0.3238	0.1552	190	-0.61	002.06	0
002	Protein, Crude (%)	0619	16.55	0.1000	16.75	0.3238	0.1552	190	-0.61	002.05	0
002	Protein, Crude (%)	0689	16.55	0.1000	16.75	0.3238	0.1552	190	-0.61	002.05	0
002	Protein, Crude (%)	0955	16.55	0.1000	16.75	0.3238	0.1552	190	-0.61	002.06	0
002	Protein, Crude (%)	0968	16.56	0.0060	16.75	0.3238	0.1552	190	-0.57	002.06	0
002	Protein, Crude (%)	0164	16.57	0.1300	16.75	0.3238	0.1552	190	-0.57	002.01	0
002	Protein, Crude (%)	0848	16.58	0.0400	16.75	0.3238	0.1552	190	-0.52	002.06	0
002	Protein, Crude (%)	0541	16.59	0.2300	16.75	0.3238	0.1552	190	-0.50	002.06	0
002	Protein, Crude (%)	2109	16.59	0.0200	16.75	0.3238	0.1552	190	-0.49	002.05	0
002	Protein, Crude (%)	0652	16.60	0.2000	16.75	0.3238	0.1552	190	-0.46	002.01	0
002	Protein, Crude (%)	0674	16.60	0.0000	16.75	0.3238	0.1552	190	-0.46	002.05	0
002	Protein, Crude (%)	0941	16.60	0.0600	16.75	0.3238	0.1552	190	-0.46	002.05	0
002	Protein, Crude (%)	0006	16.61	0.0900	16.75	0.3238	0.1552	190	-0.44	002.06	0
002	Protein, Crude (%)	2193	16.61	0.3900	16.75	0.3238	0.1552	190	-0.44	002.01	0
002	Protein, Crude (%)	0852	16.61	0.0400	16.75	0.3238	0.1552	190	-0.43	002.11	0
002	Protein, Crude (%)	0504	16.63	0.1700	16.75	0.3238	0.1552	190	-0.38	002.06	0
002	Protein, Crude (%)	0098	16.63	0.0100	16.75	0.3238	0.1552	190	-0.38	002.08	0
002	Protein, Crude (%)	2081	16.63	0.2000	16.75	0.3238	0.1552	190	-0.36	002.06	0
002	Protein, Crude (%)	0939	16.64	0.0100	16.75	0.3238	0.1552	190	-0.35	002.05	0
002	Protein, Crude (%)	0019	16.64	0.3900	16.75	0.3238	0.1552	190	-0.35	002.06	0
002	Protein, Crude (%)	0893	16.64	0.0400	16.75	0.3238	0.1552	190	-0.33	002.05	0
002	Protein, Crude (%)	0638	16.65	0.1100	16.75	0.3238	0.1552	190	-0.32	002.04	0
002	Protein, Crude (%)	0626	16.65	0.0100	16.75	0.3238	0.1552	190	-0.32	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein, Crude (%)	0651	16.65	0.0830	16.75	0.3238	0.1552	190	-0.31	002.05	0
002	Protein, Crude (%)	0265	16.65	0.1000	16.75	0.3238	0.1552	190	-0.30	002.06	0
002	Protein, Crude (%)	0956	16.65	0.3000	16.75	0.3238	0.1552	190	-0.30	002.06	0
002	Protein, Crude (%)	0123	16.67	0.0300	16.75	0.3238	0.1552	190	-0.26	002.06	0
002	Protein, Crude (%)	0066	16.67	0.0600	16.75	0.3238	0.1552	190	-0.24	002.02	0
002	Protein, Crude (%)	0868	16.68	0.0100	16.75	0.3238	0.1552	190	-0.23	002.05	0
002	Protein, Crude (%)	0015	16.69	0.3900	16.75	0.3238	0.1552	190	-0.19	002.06	0
002	Protein, Crude (%)	2190	16.69	0.3100	16.75	0.3238	0.1552	190	-0.19	002.06	0
002	Protein, Crude (%)	0676	16.69	0.1200	16.75	0.3238	0.1552	190	-0.18	002.06	0
002	Protein, Crude (%)	0695	16.70	0.3050	16.75	0.3238	0.1552	190	-0.16	002.06	0
002	Protein, Crude (%)	2196	16.70	0.0000	16.75	0.3238	0.1552	190	-0.15	002.01	0
002	Protein, Crude (%)	0884	16.70	0.0000	16.75	0.3238	0.1552	190	-0.15	002.06	0
002	Protein, Crude (%)	2184	16.70	0.3400	16.75	0.3238	0.1552	190	-0.15	002.06	0
002	Protein, Crude (%)	0592	16.72	0.0500	16.75	0.3238	0.1552	190	-0.10	002.06	0
002	Protein, Crude (%)	0405	16.72	0.0000	16.75	0.3238	0.1552	190	-0.09	002.04	0
002	Protein, Crude (%)	0675	16.72	0.0600	16.75	0.3238	0.1552	190	-0.09	002.06	0
002	Protein, Crude (%)	2006	16.73	0.0700	16.75	0.3238	0.1552	190	-0.07	002.05	0
002	Protein, Crude (%)	0001	16.73	0.3900	16.75	0.3238	0.1552	190	-0.07	002.06	0
002	Protein, Crude (%)	0610	16.73	0.0500	16.75	0.3238	0.1552	190	-0.07	002.06	0
002	Protein, Crude (%)	0038	16.73	0.0080	16.75	0.3238	0.1552	190	-0.07	002.06	0
002	Protein, Crude (%)	0872	16.73	0.2000	16.75	0.3238	0.1552	190	-0.06	002.06	0
002	Protein, Crude (%)	0878	16.73	0.0800	16.75	0.3238	0.1552	190	-0.06	002.05	0
002	Protein, Crude (%)	0202	16.73	0.0500	16.75	0.3238	0.1552	190	-0.05	002.06	0
002	Protein, Crude (%)	2089	16.74	0.0000	16.75	0.3238	0.1552	190	-0.03	002.06	0
002	Protein, Crude (%)	0589	16.74	0.0600	16.75	0.3238	0.1552	190	-0.03	002.06	0
002	Protein, Crude (%)	0098	16.75	0.3000	16.75	0.3238	0.1552	190	0.01	002.06	0
002	Protein, Crude (%)	0619	16.75	0.1000	16.75	0.3238	0.1552	190	0.01	002.06	0
002	Protein, Crude (%)	0407	16.76	0.1150	16.75	0.3238	0.1552	190	0.03	002.06	0
002	Protein, Crude (%)	2144	16.76	0.1400	16.75	0.3238	0.1552	190	0.04	002.05	0
002	Protein, Crude (%)	0028	16.77	0.0100	16.75	0.3238	0.1552	190	0.05	002.06	0
002	Protein, Crude (%)	2058	16.77	0.0500	16.75	0.3238	0.1552	190	0.05	002.06	0
002	Protein, Crude (%)	0226	16.77	0.0600	16.75	0.3238	0.1552	190	0.07	002.06	0
002	Protein, Crude (%)	0610	16.78	0.0500	16.75	0.3238	0.1552	190	0.08	002.01	0
002	Protein, Crude (%)	0187	16.78	0.0500	16.75	0.3238	0.1552	190	0.08	002.04	0
002	Protein, Crude (%)	0504	16.78	0.0500	16.75	0.3238	0.1552	190	0.08	002.04	0
002	Protein, Crude (%)	0868	16.78	0.0100	16.75	0.3238	0.1552	190	0.08	002.06	0
002	Protein, Crude (%)	0164	16.78	0.0600	16.75	0.3238	0.1552	190	0.10	002.06	0
002	Protein, Crude (%)	2150	16.79	0.3900	16.75	0.3238	0.1552	190	0.11	002.06	0
002	Protein, Crude (%)	0590	16.79	0.2800	16.75	0.3238	0.1552	190	0.13	002.06	0
002	Protein, Crude (%)	2222	16.79	0.1600	16.75	0.3238	0.1552	190	0.13	002.06	0
002	Protein, Crude (%)	2022	16.80	0.0100	16.75	0.3238	0.1552	190	0.14	002.05	0
002	Protein, Crude (%)	0010	16.80	0.2000	16.75	0.3238	0.1552	190	0.16	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein, Crude (%)	2129	16.80	0.0600	16.75	0.3238	0.1552	190	0.16	002.06	0
002	Protein, Crude (%)	0278	16.80	0.0000	16.75	0.3238	0.1552	190	0.16	002.06	0
002	Protein, Crude (%)	0910	16.80	0.0000	16.75	0.3238	0.1552	190	0.16	002.06	0
002	Protein, Crude (%)	0870	16.81	0.1100	16.75	0.3238	0.1552	190	0.18	002.06	0
002	Protein, Crude (%)	0357	16.81	0.1250	16.75	0.3238	0.1552	190	0.20	002.06	0
002	Protein, Crude (%)	0897	16.82	0.0100	16.75	0.3238	0.1552	190	0.21	002.05	0
002	Protein, Crude (%)	0205	16.82	0.1100	16.75	0.3238	0.1552	190	0.21	002.06	0
002	Protein, Crude (%)	0948	16.82	0.0200	16.75	0.3238	0.1552	190	0.22	002.05	0
002	Protein, Crude (%)	0622	16.84	0.4800	16.75	0.3238	0.1552	190	0.28	002.05	0
002	Protein, Crude (%)	0017	16.85	0.3000	16.75	0.3238	0.1552	190	0.31	002.06	0
002	Protein, Crude (%)	0074	16.85	0.3200	16.75	0.3238	0.1552	190	0.31	002.06	0
002	Protein, Crude (%)	0148	16.86	0.2300	16.75	0.3238	0.1552	190	0.33	002.06	0
002	Protein, Crude (%)	2053	16.86	0.2700	16.75	0.3238	0.1552	190	0.33	002.06	0
002	Protein, Crude (%)	0358	16.86	0.6700	16.75	0.3238	0.1552	190	0.33	002.06	0
002	Protein, Crude (%)	0968	16.86	0.0300	16.75	0.3238	0.1552	190	0.34	002.01	0
002	Protein, Crude (%)	0693	16.86	0.3490	16.75	0.3238	0.1552	190	0.36	002.06	0
002	Protein, Crude (%)	0309	16.87	0.0260	16.75	0.3238	0.1552	190	0.38	002.06	0
002	Protein, Crude (%)	2168	16.87	0.0800	16.75	0.3238	0.1552	190	0.38	002.06	0
002	Protein, Crude (%)	0857	16.88	0.1100	16.75	0.3238	0.1552	190	0.39	002.06	0
002	Protein, Crude (%)	0035	16.88	0.1000	16.75	0.3238	0.1552	190	0.41	002.06	0
002	Protein, Crude (%)	0089	16.89	0.0100	16.75	0.3238	0.1552	190	0.42	002.06	0
002	Protein, Crude (%)	0009	16.90	0.2200	16.75	0.3238	0.1552	190	0.47	002.06	0
002	Protein, Crude (%)	0298	16.90	0.2000	16.75	0.3238	0.1552	190	0.47	002.06	0
002	Protein, Crude (%)	0142	16.93	0.0100	16.75	0.3238	0.1552	190	0.55	002.06	0
002	Protein, Crude (%)	0004	16.93	0.0400	16.75	0.3238	0.1552	190	0.56	002.06	0
002	Protein, Crude (%)	0036	16.93	0.1956	16.75	0.3238	0.1552	190	0.58	002.02	0
002	Protein, Crude (%)	0425	16.95	0.1300	16.75	0.3238	0.1552	190	0.61	002.06	0
002	Protein, Crude (%)	0033	16.95	0.5000	16.75	0.3238	0.1552	190	0.62	002.06	0
002	Protein, Crude (%)	0948	16.95	0.0400	16.75	0.3238	0.1552	190	0.62	002.06	0
002	Protein, Crude (%)	0108	16.96	0.4900	16.75	0.3238	0.1552	190	0.64	002.06	0
002	Protein, Crude (%)	0037	16.97	0.5700	16.75	0.3238	0.1552	190	0.67	002.06	0
002	Protein, Crude (%)	0843	16.98	0.0200	16.75	0.3238	0.1552	190	0.72	002.06	0
002	Protein, Crude (%)	2057	16.99	0.0210	16.75	0.3238	0.1552	190	0.73	002.06	0
002	Protein, Crude (%)	0353	16.99	0.1800	16.75	0.3238	0.1552	190	0.75	002.06	0
002	Protein, Crude (%)	2200	17.00	0.4000	16.75	0.3238	0.1552	190	0.78	002.05	0
002	Protein, Crude (%)	0366	17.00	0.6000	16.75	0.3238	0.1552	190	0.78	002.06	0
002	Protein, Crude (%)	2242	17.00	0.2000	16.75	0.3238	0.1552	190	0.78	002.06	0
002	Protein, Crude (%)	0939	17.02	0.1400	16.75	0.3238	0.1552	190	0.84	002.06	0
002	Protein, Crude (%)	0571	17.02	0.1600	16.75	0.3238	0.1552	190	0.84	002.06	0
002	Protein, Crude (%)	0242	17.03	0.0500	16.75	0.3238	0.1552	190	0.86	002.06	0
002	Protein, Crude (%)	2144	17.03	0.0400	16.75	0.3238	0.1552	190	0.87	002.06	0
002	Protein, Crude (%)	0036	17.05	0.0100	16.75	0.3238	0.1552	190	0.92	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein, Crude (%)	2018	17.05	0.0100	16.75	0.3238	0.1552	190	0.92	002.06	0
002	Protein, Crude (%)	0045	17.05	0.1000	16.75	0.3238	0.1552	190	0.93	002.06	0
002	Protein, Crude (%)	0186	17.05	0.1000	16.75	0.3238	0.1552	190	0.93	002.06	0
002	Protein, Crude (%)	0596	17.07	0.0500	16.75	0.3238	0.1552	190	0.98	002.06	0
002	Protein, Crude (%)	0190	17.08	0.1700	16.75	0.3238	0.1552	190	1.01	002.06	0
002	Protein, Crude (%)	0574	17.08	0.0500	16.75	0.3238	0.1552	190	1.01	002.06	0
002	Protein, Crude (%)	0656	17.08	0.3100	16.75	0.3238	0.1552	190	1.01	002.06	0
002	Protein, Crude (%)	0859	17.08	0.0160	16.75	0.3238	0.1552	190	1.02	002.06	0
002	Protein, Crude (%)	0003	17.08	0.2000	16.75	0.3238	0.1552	190	1.03	002.06	0
002	Protein, Crude (%)	0042	17.09	0.0200	16.75	0.3238	0.1552	190	1.06	002.06	0
002	Protein, Crude (%)	0646	17.09	0.1200	16.75	0.3238	0.1552	190	1.06	002.06	0
002	Protein, Crude (%)	0511	17.10	0.0500	16.75	0.3238	0.1552	190	1.07	002.06	0
002	Protein, Crude (%)	0034	17.10	0.1590	16.75	0.3238	0.1552	190	1.08	002.06	0
002	Protein, Crude (%)	0682	17.10	0.0000	16.75	0.3238	0.1552	190	1.09	002.06	0
002	Protein, Crude (%)	0918	17.11	0.3500	16.75	0.3238	0.1552	190	1.10	002.06	0
002	Protein, Crude (%)	2126	17.11	0.0400	16.75	0.3238	0.1552	190	1.12	002.06	0
002	Protein, Crude (%)	0687	17.15	0.7100	16.75	0.3238	0.1552	190	1.23	002.06	0
002	Protein, Crude (%)	0083	17.15	0.1000	16.75	0.3238	0.1552	190	1.24	002.06	0
002	Protein, Crude (%)	2076	17.16	0.2840	16.75	0.3238	0.1552	190	1.28	002.06	0
002	Protein, Crude (%)	2033	17.17	0.1800	16.75	0.3238	0.1552	190	1.30	002.06	0
002	Protein, Crude (%)	0229	17.19	0.2400	16.75	0.3238	0.1552	190	1.36	002.06	0
002	Protein, Crude (%)	0520	17.20	0.2000	16.75	0.3238	0.1552	190	1.40	002.06	0
002	Protein, Crude (%)	0726	17.21	0.0300	16.75	0.3238	0.1552	190	1.41	002.06	0
002	Protein, Crude (%)	0263	17.21	0.0170	16.75	0.3238	0.1552	190	1.42	002.06	0
002	Protein, Crude (%)	2213	17.25	0.2500	16.75	0.3238	0.1552	190	1.53	002.06	0
002	Protein, Crude (%)	0553	17.26	0.2000	16.75	0.3238	0.1552	190	1.58	002.06	0
002	Protein, Crude (%)	0563	17.27	0.0070	16.75	0.3238	0.1552	190	1.60	002.06	0
002	Protein, Crude (%)	0876	17.27	0.2400	16.75	0.3238	0.1552	190	1.61	002.06	0
002	Protein, Crude (%)	2059	17.32	0.0969	16.75	0.3238	0.1552	190	1.77	002.06	0
002	Protein, Crude (%)	2246	17.37	0.0100	16.75	0.3238	0.1552	190	1.91	002.06	0
002	Protein, Crude (%)	0563	17.68	0.0491	16.75	0.3238	0.1552	190	2.88	002.08	0
002	Protein, Crude (%)	0618	17.74	0.1250	16.75	0.3238	0.1552	190	3.05	002.06	0
002	Protein, Crude (%)	0953	18.36	0.2600	16.75	0.3238	0.1552	190	4.98	002.06	0
002	Protein, Crude (%)	2128	18.78	0.1760	16.75	0.3238	0.1552	190	6.26	002.11	0
002	Protein, Crude (%)	0853	19.15	0.1000	16.75	0.3238	0.1552	190	7.42	002.06	0
002	Protein, Crude (%)	0720	19.49	0.2100	16.75	0.3238	0.1552	190	8.45	002.11	0
002	Protein, Crude (%)	0612	19.53	0.5700	16.75	0.3238	0.1552	190	8.58	002.06	0
002	Protein, Crude (%)	0553	15.90	0.8300	16.75	0.3238	0.1552	190	-2.63	002.11	1
002	Protein, Crude (%)	0643	16.76	0.8900	16.75	0.3238	0.1552	190	0.02	002.99	1
002	Protein, Crude (%)	0049	16.87	0.7600	16.75	0.3238	0.1552	190	0.38	002.06	1
002	Protein, Crude (%)	0297	23.04	0.1200	16.75	0.3238	0.1552	190	19.43	002.11	2
002	Protein, Crude (%)	0297	24.34	0.1000	16.75	0.3238	0.1552	190	23.45	002.06	2

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat, Crude (%)	2144	2.190	0.0400	3.360	0.3098	0.0880	145	-3.78	003.14	0
003	Fat, Crude (%)	0263	2.197	0.0100	3.360	0.3098	0.0880	145	-3.76	003.06	0
003	Fat, Crude (%)	0504	2.230	0.3400	3.360	0.3098	0.0880	145	-3.65	003.01	0
003	Fat, Crude (%)	0019	2.645	0.0100	3.360	0.3098	0.0880	145	-2.31	003.14	0
003	Fat, Crude (%)	0843	2.795	0.0300	3.360	0.3098	0.0880	145	-1.82	003.14	0
003	Fat, Crude (%)	0083	2.860	0.0800	3.360	0.3098	0.0880	145	-1.62	003.14	0
003	Fat, Crude (%)	0037	2.865	0.1700	3.360	0.3098	0.0880	145	-1.60	003.14	0
003	Fat, Crude (%)	0651	2.868	0.1500	3.360	0.3098	0.0880	145	-1.59	003.10	0
003	Fat, Crude (%)	2188	2.905	0.0500	3.360	0.3098	0.0880	145	-1.47	003.10	0
003	Fat, Crude (%)	0574	2.915	0.0500	3.360	0.3098	0.0880	145	-1.44	003.14	0
003	Fat, Crude (%)	0242	2.940	0.0200	3.360	0.3098	0.0880	145	-1.36	003.10	0
003	Fat, Crude (%)	0581	2.960	0.0200	3.360	0.3098	0.0880	145	-1.29	003.14	0
003	Fat, Crude (%)	2073	2.990	0.0200	3.360	0.3098	0.0880	145	-1.20	003.10	0
003	Fat, Crude (%)	0885	3.011	0.1257	3.360	0.3098	0.0880	145	-1.13	003.00	0
003	Fat, Crude (%)	0358	3.020	0.2400	3.360	0.3098	0.0880	145	-1.10	003.14	0
003	Fat, Crude (%)	0968	3.027	0.0300	3.360	0.3098	0.0880	145	-1.08	003.14	0
003	Fat, Crude (%)	2193	3.030	0.0200	3.360	0.3098	0.0880	145	-1.07	003.10	0
003	Fat, Crude (%)	0876	3.040	0.4000	3.360	0.3098	0.0880	145	-1.03	003.14	0
003	Fat, Crude (%)	0553	3.045	0.1900	3.360	0.3098	0.0880	145	-1.02	003.10	0
003	Fat, Crude (%)	0144	3.050	0.0400	3.360	0.3098	0.0880	145	-1.00	003.14	0
003	Fat, Crude (%)	0045	3.055	0.0900	3.360	0.3098	0.0880	145	-0.99	003.10	0
003	Fat, Crude (%)	2246	3.055	0.0300	3.360	0.3098	0.0880	145	-0.99	003.13	0
003	Fat, Crude (%)	2089	3.055	0.1700	3.360	0.3098	0.0880	145	-0.99	003.13	0
003	Fat, Crude (%)	2076	3.061	0.3150	3.360	0.3098	0.0880	145	-0.97	003.14	0
003	Fat, Crude (%)	0229	3.064	0.0070	3.360	0.3098	0.0880	145	-0.96	003.14	0
003	Fat, Crude (%)	2258	3.085	0.0500	3.360	0.3098	0.0880	145	-0.89	003.00	0
003	Fat, Crude (%)	0003	3.085	0.0500	3.360	0.3098	0.0880	145	-0.89	003.14	0
003	Fat, Crude (%)	0298	3.090	0.0200	3.360	0.3098	0.0880	145	-0.87	003.14	0
003	Fat, Crude (%)	0520	3.095	0.0700	3.360	0.3098	0.0880	145	-0.86	003.14	0
003	Fat, Crude (%)	2152	3.100	0.0200	3.360	0.3098	0.0880	145	-0.84	003.12	0
003	Fat, Crude (%)	0693	3.103	0.1270	3.360	0.3098	0.0880	145	-0.83	003.10	0
003	Fat, Crude (%)	0186	3.115	0.0300	3.360	0.3098	0.0880	145	-0.79	003.14	0
003	Fat, Crude (%)	0036	3.122	0.1038	3.360	0.3098	0.0880	145	-0.77	003.10	0
003	Fat, Crude (%)	0536	3.125	0.0100	3.360	0.3098	0.0880	145	-0.76	003.99	0
003	Fat, Crude (%)	0682	3.130	0.0000	3.360	0.3098	0.0880	145	-0.74	003.14	0
003	Fat, Crude (%)	0870	3.136	0.0098	3.360	0.3098	0.0880	145	-0.72	003.10	0
003	Fat, Crude (%)	2222	3.140	0.0600	3.360	0.3098	0.0880	145	-0.71	003.10	0
003	Fat, Crude (%)	0511	3.140	0.0200	3.360	0.3098	0.0880	145	-0.71	003.14	0
003	Fat, Crude (%)	0956	3.150	0.1000	3.360	0.3098	0.0880	145	-0.68	003.14	0
003	Fat, Crude (%)	0089	3.155	0.0100	3.360	0.3098	0.0880	145	-0.66	003.14	0
003	Fat, Crude (%)	2022	3.160	0.0400	3.360	0.3098	0.0880	145	-0.65	003.10	0
003	Fat, Crude (%)	0265	3.160	0.0200	3.360	0.3098	0.0880	145	-0.65	003.14	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat, Crude (%)	0939	3.170	0.0400	3.360	0.3098	0.0880	145	-0.61	003.99	0
003	Fat, Crude (%)	0100	3.175	0.1900	3.360	0.3098	0.0880	145	-0.60	003.10	0
003	Fat, Crude (%)	0893	3.180	0.0000	3.360	0.3098	0.0880	145	-0.58	003.12	0
003	Fat, Crude (%)	0366	3.200	0.2000	3.360	0.3098	0.0880	145	-0.52	003.10	0
003	Fat, Crude (%)	0720	3.210	0.0400	3.360	0.3098	0.0880	145	-0.49	003.10	0
003	Fat, Crude (%)	0098	3.215	0.1500	3.360	0.3098	0.0880	145	-0.47	003.10	0
003	Fat, Crude (%)	0868	3.215	0.0300	3.360	0.3098	0.0880	145	-0.47	003.14	0
003	Fat, Crude (%)	0062	3.226	0.0900	3.360	0.3098	0.0880	145	-0.43	003.10	0
003	Fat, Crude (%)	2155	3.228	0.0690	3.360	0.3098	0.0880	145	-0.43	003.10	0
003	Fat, Crude (%)	0541	3.230	0.1000	3.360	0.3098	0.0880	145	-0.42	003.10	0
003	Fat, Crude (%)	2057	3.232	0.0746	3.360	0.3098	0.0880	145	-0.41	003.14	0
003	Fat, Crude (%)	0278	3.250	0.1000	3.360	0.3098	0.0880	145	-0.36	003.14	0
003	Fat, Crude (%)	0357	3.250	0.1000	3.360	0.3098	0.0880	145	-0.36	003.14	0
003	Fat, Crude (%)	0035	3.275	0.0100	3.360	0.3098	0.0880	145	-0.28	003.14	0
003	Fat, Crude (%)	0505	3.275	0.1900	3.360	0.3098	0.0880	145	-0.28	003.14	0
003	Fat, Crude (%)	0202	3.278	0.1450	3.360	0.3098	0.0880	145	-0.27	003.14	0
003	Fat, Crude (%)	0955	3.280	0.0200	3.360	0.3098	0.0880	145	-0.26	003.14	0
003	Fat, Crude (%)	0878	3.285	0.0300	3.360	0.3098	0.0880	145	-0.24	003.14	0
003	Fat, Crude (%)	2190	3.290	0.1200	3.360	0.3098	0.0880	145	-0.23	003.00	0
003	Fat, Crude (%)	0644	3.293	0.0380	3.360	0.3098	0.0880	145	-0.22	003.06	0
003	Fat, Crude (%)	0004	3.295	0.1500	3.360	0.3098	0.0880	145	-0.21	003.14	0
003	Fat, Crude (%)	2199	3.300	0.2000	3.360	0.3098	0.0880	145	-0.19	003.12	0
003	Fat, Crude (%)	0033	3.300	0.0200	3.360	0.3098	0.0880	145	-0.19	003.13	0
003	Fat, Crude (%)	2184	3.310	0.0200	3.360	0.3098	0.0880	145	-0.16	003.10	0
003	Fat, Crude (%)	0123	3.315	0.0100	3.360	0.3098	0.0880	145	-0.15	003.13	0
003	Fat, Crude (%)	0049	3.315	0.3900	3.360	0.3098	0.0880	145	-0.15	003.14	0
003	Fat, Crude (%)	0066	3.330	0.0600	3.360	0.3098	0.0880	145	-0.10	003.14	0
003	Fat, Crude (%)	0529	3.330	0.2000	3.360	0.3098	0.0880	145	-0.10	003.14	0
003	Fat, Crude (%)	0510	3.350	0.1000	3.360	0.3098	0.0880	145	-0.03	003.09	0
003	Fat, Crude (%)	0897	3.350	0.0000	3.360	0.3098	0.0880	145	-0.03	003.10	0
003	Fat, Crude (%)	0407	3.350	0.1854	3.360	0.3098	0.0880	145	-0.03	003.14	0
003	Fat, Crude (%)	0889	3.360	0.1200	3.360	0.3098	0.0880	145	0.00	003.10	0
003	Fat, Crude (%)	2196	3.360	0.0000	3.360	0.3098	0.0880	145	0.00	003.10	0
003	Fat, Crude (%)	0948	3.360	0.1200	3.360	0.3098	0.0880	145	0.00	003.14	0
003	Fat, Crude (%)	0309	3.362	0.2033	3.360	0.3098	0.0880	145	0.01	003.00	0
003	Fat, Crude (%)	0108	3.370	0.0400	3.360	0.3098	0.0880	145	0.03	003.14	0
003	Fat, Crude (%)	0948	3.375	0.1300	3.360	0.3098	0.0880	145	0.05	003.09	0
003	Fat, Crude (%)	0205	3.379	0.0060	3.360	0.3098	0.0880	145	0.06	003.13	0
003	Fat, Crude (%)	0034	3.380	0.3000	3.360	0.3098	0.0880	145	0.06	003.14	0
003	Fat, Crude (%)	0169	3.385	0.0500	3.360	0.3098	0.0880	145	0.08	003.06	0
003	Fat, Crude (%)	2200	3.400	0.0000	3.360	0.3098	0.0880	145	0.13	003.12	0
003	Fat, Crude (%)	0148	3.400	0.0200	3.360	0.3098	0.0880	145	0.13	003.10	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat, Crude (%)	0618	3.401	0.0240	3.360	0.3098	0.0880	145	0.13	003.10	0
003	Fat, Crude (%)	0074	3.405	0.1700	3.360	0.3098	0.0880	145	0.14	003.14	0
003	Fat, Crude (%)	0683	3.410	0.2000	3.360	0.3098	0.0880	145	0.16	003.14	0
003	Fat, Crude (%)	0695	3.420	0.1200	3.360	0.3098	0.0880	145	0.19	003.10	0
003	Fat, Crude (%)	0229	3.428	0.0130	3.360	0.3098	0.0880	145	0.22	003.06	0
003	Fat, Crude (%)	0294	3.450	0.1200	3.360	0.3098	0.0880	145	0.29	003.06	0
003	Fat, Crude (%)	0042	3.455	0.1700	3.360	0.3098	0.0880	145	0.31	003.10	0
003	Fat, Crude (%)	0226	3.465	0.0100	3.360	0.3098	0.0880	145	0.34	003.09	0
003	Fat, Crude (%)	2181	3.470	0.1000	3.360	0.3098	0.0880	145	0.35	003.06	0
003	Fat, Crude (%)	0004	3.470	0.1200	3.360	0.3098	0.0880	145	0.35	003.09	0
003	Fat, Crude (%)	0646	3.470	0.1400	3.360	0.3098	0.0880	145	0.35	003.13	0
003	Fat, Crude (%)	0171	3.495	0.1300	3.360	0.3098	0.0880	145	0.43	003.12	0
003	Fat, Crude (%)	0689	3.500	0.0000	3.360	0.3098	0.0880	145	0.45	003.10	0
003	Fat, Crude (%)	0720	3.500	0.0400	3.360	0.3098	0.0880	145	0.45	003.11	0
003	Fat, Crude (%)	0849	3.513	0.0250	3.360	0.3098	0.0880	145	0.49	003.06	0
003	Fat, Crude (%)	0563	3.515	0.0117	3.360	0.3098	0.0880	145	0.50	003.14	0
003	Fat, Crude (%)	2192	3.520	0.1200	3.360	0.3098	0.0880	145	0.52	003.00	0
003	Fat, Crude (%)	0009	3.520	0.1400	3.360	0.3098	0.0880	145	0.52	003.14	0
003	Fat, Crude (%)	0675	3.530	0.1200	3.360	0.3098	0.0880	145	0.55	003.09	0
003	Fat, Crude (%)	2062	3.545	0.0461	3.360	0.3098	0.0880	145	0.60	003.12	0
003	Fat, Crude (%)	0353	3.550	0.2600	3.360	0.3098	0.0880	145	0.61	003.09	0
003	Fat, Crude (%)	0638	3.550	0.2000	3.360	0.3098	0.0880	145	0.61	003.09	0
003	Fat, Crude (%)	0098	3.555	0.0100	3.360	0.3098	0.0880	145	0.63	003.09	0
003	Fat, Crude (%)	0619	3.555	0.1300	3.360	0.3098	0.0880	145	0.63	003.14	0
003	Fat, Crude (%)	0598	3.560	0.1600	3.360	0.3098	0.0880	145	0.64	003.14	0
003	Fat, Crude (%)	0345	3.565	0.0900	3.360	0.3098	0.0880	145	0.66	003.00	0
003	Fat, Crude (%)	2109	3.580	0.0600	3.360	0.3098	0.0880	145	0.71	003.00	0
003	Fat, Crude (%)	0910	3.585	0.0300	3.360	0.3098	0.0880	145	0.73	003.06	0
003	Fat, Crude (%)	0870	3.587	0.0203	3.360	0.3098	0.0880	145	0.73	003.09	0
003	Fat, Crude (%)	0148	3.595	0.0300	3.360	0.3098	0.0880	145	0.76	003.14	0
003	Fat, Crude (%)	0870	3.598	0.0473	3.360	0.3098	0.0880	145	0.77	003.13	0
003	Fat, Crude (%)	0674	3.600	0.0000	3.360	0.3098	0.0880	145	0.77	003.09	0
003	Fat, Crude (%)	2006	3.600	0.2800	3.360	0.3098	0.0880	145	0.77	003.13	0
003	Fat, Crude (%)	0848	3.615	0.0300	3.360	0.3098	0.0880	145	0.82	003.00	0
003	Fat, Crude (%)	0918	3.635	0.2500	3.360	0.3098	0.0880	145	0.89	003.06	0
003	Fat, Crude (%)	0689	3.650	0.1000	3.360	0.3098	0.0880	145	0.93	003.06	0
003	Fat, Crude (%)	0726	3.660	0.0000	3.360	0.3098	0.0880	145	0.97	003.14	0
003	Fat, Crude (%)	2128	3.679	0.0095	3.360	0.3098	0.0880	145	1.03	003.11	0
003	Fat, Crude (%)	0682	3.680	0.0000	3.360	0.3098	0.0880	145	1.03	003.06	0
003	Fat, Crude (%)	0001	3.681	0.0010	3.360	0.3098	0.0880	145	1.03	003.14	0
003	Fat, Crude (%)	0676	3.695	0.0900	3.360	0.3098	0.0880	145	1.08	003.10	0
003	Fat, Crude (%)	2058	3.695	0.0100	3.360	0.3098	0.0880	145	1.08	003.12	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat, Crude (%)	2058	3.725	0.0500	3.360	0.3098	0.0880	145	1.18	003.99	0
003	Fat, Crude (%)	0035	3.765	0.1300	3.360	0.3098	0.0880	145	1.31	003.00	0
003	Fat, Crude (%)	2053	3.780	0.0600	3.360	0.3098	0.0880	145	1.35	003.00	0
003	Fat, Crude (%)	0405	3.790	0.0000	3.360	0.3098	0.0880	145	1.39	003.11	0
003	Fat, Crude (%)	0941	3.795	0.1100	3.360	0.3098	0.0880	145	1.40	003.99	0
003	Fat, Crude (%)	0425	3.810	0.0200	3.360	0.3098	0.0880	145	1.45	003.06	0
003	Fat, Crude (%)	0656	3.825	0.4700	3.360	0.3098	0.0880	145	1.50	003.06	0
003	Fat, Crude (%)	0164	3.840	0.0400	3.360	0.3098	0.0880	145	1.55	003.06	0
003	Fat, Crude (%)	0017	3.850	0.1000	3.360	0.3098	0.0880	145	1.58	003.00	0
003	Fat, Crude (%)	0590	3.870	0.0200	3.360	0.3098	0.0880	145	1.65	003.09	0
003	Fat, Crude (%)	2126	3.900	0.0600	3.360	0.3098	0.0880	145	1.74	003.00	0
003	Fat, Crude (%)	0190	3.920	0.0600	3.360	0.3098	0.0880	145	1.81	003.14	0
003	Fat, Crude (%)	0553	3.945	0.0700	3.360	0.3098	0.0880	145	1.89	003.11	0
003	Fat, Crude (%)	2018	4.035	0.0500	3.360	0.3098	0.0880	145	2.18	003.10	0
003	Fat, Crude (%)	0852	4.065	0.0300	3.360	0.3098	0.0880	145	2.27	003.11	0
003	Fat, Crude (%)	2144	4.165	0.0500	3.360	0.3098	0.0880	145	2.60	003.00	0
003	Fat, Crude (%)	0011	4.300	0.0000	3.360	0.3098	0.0880	145	3.03	003.11	0
003	Fat, Crude (%)	0884	4.550	0.1000	3.360	0.3098	0.0880	145	3.84	003.06	0
003	Fat, Crude (%)	0297	5.790	0.2600	3.360	0.3098	0.0880	145	7.84	003.06	0
003	Fat, Crude (%)	2129	3.005	0.4900	3.360	0.3098	0.0880	145	-1.15	003.14	1
003	Fat, Crude (%)	0853	3.150	0.5000	3.360	0.3098	0.0880	145	-0.68	003.14	1
003	Fat, Crude (%)	0687	3.500	0.6000	3.360	0.3098	0.0880	145	0.45	003.06	1
003	Fat, Crude (%)	0098	4.250	0.5000	3.360	0.3098	0.0880	145	2.87	003.13	1
003	Fat, Crude (%)	0297	9.140	0.0000	3.360	0.3098	0.0880	145	18.66	003.11	2
004	Fiber, Crude (%)	0852	9.415	0.0500	13.41	0.8105	0.2308	114	-4.92	004.11	0
004	Fiber, Crude (%)	2089	10.46	0.1800	13.41	0.8105	0.2308	114	-3.64	004.03	0
004	Fiber, Crude (%)	0720	11.20	0.2000	13.41	0.8105	0.2308	114	-2.72	004.11	0
004	Fiber, Crude (%)	0366	11.75	0.5000	13.41	0.8105	0.2308	114	-2.04	004.01	0
004	Fiber, Crude (%)	0405	11.99	0.0000	13.41	0.8105	0.2308	114	-1.75	004.11	0
004	Fiber, Crude (%)	0682	12.00	0.0000	13.41	0.8105	0.2308	114	-1.74	004.07	0
004	Fiber, Crude (%)	0042	12.06	0.1600	13.41	0.8105	0.2308	114	-1.66	004.07	0
004	Fiber, Crude (%)	0004	12.20	0.5300	13.41	0.8105	0.2308	114	-1.49	004.07	0
004	Fiber, Crude (%)	0278	12.20	0.2000	13.41	0.8105	0.2308	114	-1.49	004.07	0
004	Fiber, Crude (%)	2144	12.32	0.4000	13.41	0.8105	0.2308	114	-1.34	004.99	0
004	Fiber, Crude (%)	0683	12.38	0.0300	13.41	0.8105	0.2308	114	-1.27	004.07	0
004	Fiber, Crude (%)	2128	12.38	0.1750	13.41	0.8105	0.2308	114	-1.26	004.11	0
004	Fiber, Crude (%)	0910	12.45	0.1000	13.41	0.8105	0.2308	114	-1.18	004.07	0
004	Fiber, Crude (%)	0590	12.50	0.0300	13.41	0.8105	0.2308	114	-1.12	004.06	0
004	Fiber, Crude (%)	0265	12.55	0.1000	13.41	0.8105	0.2308	114	-1.06	004.07	0
004	Fiber, Crude (%)	0956	12.60	0.8000	13.41	0.8105	0.2308	114	-1.00	004.07	0
004	Fiber, Crude (%)	0619	12.65	0.1000	13.41	0.8105	0.2308	114	-0.93	004.07	0
004	Fiber, Crude (%)	0884	12.70	0.2000	13.41	0.8105	0.2308	114	-0.87	004.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber, Crude (%)	2146	12.72	0.1400	13.41	0.8105	0.2308	114	-0.85	004.99	0
004	Fiber, Crude (%)	0144	12.73	0.0900	13.41	0.8105	0.2308	114	-0.84	004.07	0
004	Fiber, Crude (%)	0646	12.74	0.0300	13.41	0.8105	0.2308	114	-0.83	004.07	0
004	Fiber, Crude (%)	2181	12.74	0.0000	13.41	0.8105	0.2308	114	-0.82	004.07	0
004	Fiber, Crude (%)	0345	12.80	0.2000	13.41	0.8105	0.2308	114	-0.75	004.00	0
004	Fiber, Crude (%)	2190	12.81	0.5800	13.41	0.8105	0.2308	114	-0.74	004.07	0
004	Fiber, Crude (%)	2062	12.84	0.0959	13.41	0.8105	0.2308	114	-0.70	004.06	0
004	Fiber, Crude (%)	0590	12.87	0.2300	13.41	0.8105	0.2308	114	-0.67	004.07	0
004	Fiber, Crude (%)	0123	12.88	0.0300	13.41	0.8105	0.2308	114	-0.66	004.07	0
004	Fiber, Crude (%)	2022	12.92	0.0100	13.41	0.8105	0.2308	114	-0.61	004.06	0
004	Fiber, Crude (%)	0098	12.93	0.0200	13.41	0.8105	0.2308	114	-0.59	004.07	0
004	Fiber, Crude (%)	0425	12.95	0.1000	13.41	0.8105	0.2308	114	-0.56	004.00	0
004	Fiber, Crude (%)	0897	12.97	0.0000	13.41	0.8105	0.2308	114	-0.54	004.06	0
004	Fiber, Crude (%)	0644	12.97	0.0660	13.41	0.8105	0.2308	114	-0.53	004.07	0
004	Fiber, Crude (%)	0009	12.99	0.3800	13.41	0.8105	0.2308	114	-0.51	004.07	0
004	Fiber, Crude (%)	0202	12.99	0.4690	13.41	0.8105	0.2308	114	-0.51	004.07	0
004	Fiber, Crude (%)	0581	13.00	0.4000	13.41	0.8105	0.2308	114	-0.50	004.07	0
004	Fiber, Crude (%)	0353	13.04	0.0700	13.41	0.8105	0.2308	114	-0.46	004.03	0
004	Fiber, Crude (%)	0028	13.05	0.9000	13.41	0.8105	0.2308	114	-0.44	004.07	0
004	Fiber, Crude (%)	0510	13.05	0.1000	13.41	0.8105	0.2308	114	-0.44	004.07	0
004	Fiber, Crude (%)	0035	13.07	0.0600	13.41	0.8105	0.2308	114	-0.42	004.07	0
004	Fiber, Crude (%)	2144	13.08	0.8400	13.41	0.8105	0.2308	114	-0.40	004.07	0
004	Fiber, Crude (%)	0968	13.14	0.0700	13.41	0.8105	0.2308	114	-0.33	004.07	0
004	Fiber, Crude (%)	0190	13.15	0.1000	13.41	0.8105	0.2308	114	-0.32	004.07	0
004	Fiber, Crude (%)	0520	13.15	0.7000	13.41	0.8105	0.2308	114	-0.32	004.07	0
004	Fiber, Crude (%)	0263	13.16	0.0327	13.41	0.8105	0.2308	114	-0.30	004.07	0
004	Fiber, Crude (%)	0297	13.19	0.2300	13.41	0.8105	0.2308	114	-0.27	004.07	0
004	Fiber, Crude (%)	0098	13.20	0.5500	13.41	0.8105	0.2308	114	-0.26	004.06	0
004	Fiber, Crude (%)	0695	13.20	0.3870	13.41	0.8105	0.2308	114	-0.26	004.00	0
004	Fiber, Crude (%)	2033	13.20	0.0000	13.41	0.8105	0.2308	114	-0.25	004.07	0
004	Fiber, Crude (%)	0610	13.20	0.2000	13.41	0.8105	0.2308	114	-0.25	004.99	0
004	Fiber, Crude (%)	0205	13.25	0.0700	13.41	0.8105	0.2308	114	-0.20	004.06	0
004	Fiber, Crude (%)	0708	13.27	0.3300	13.41	0.8105	0.2308	114	-0.17	004.07	0
004	Fiber, Crude (%)	2188	13.27	0.0100	13.41	0.8105	0.2308	114	-0.17	004.07	0
004	Fiber, Crude (%)	0169	13.27	0.2800	13.41	0.8105	0.2308	114	-0.17	004.00	0
004	Fiber, Crude (%)	0726	13.29	0.0900	13.41	0.8105	0.2308	114	-0.15	004.07	0
004	Fiber, Crude (%)	0939	13.29	0.1700	13.41	0.8105	0.2308	114	-0.15	004.07	0
004	Fiber, Crude (%)	0968	13.29	0.0250	13.41	0.8105	0.2308	114	-0.14	004.06	0
004	Fiber, Crude (%)	2200	13.30	0.6000	13.41	0.8105	0.2308	114	-0.13	004.06	0
004	Fiber, Crude (%)	0089	13.32	0.0400	13.41	0.8105	0.2308	114	-0.11	004.07	0
004	Fiber, Crude (%)	0083	13.33	0.2500	13.41	0.8105	0.2308	114	-0.10	004.07	0
004	Fiber, Crude (%)	0868	13.39	0.0900	13.41	0.8105	0.2308	114	-0.03	004.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber, Crude (%)	0638	13.40	0.4000	13.41	0.8105	0.2308	114	-0.01	004.06	0
004	Fiber, Crude (%)	0643	13.40	0.1400	13.41	0.8105	0.2308	114	-0.01	004.07	0
004	Fiber, Crude (%)	2058	13.41	0.4100	13.41	0.8105	0.2308	114	0.00	004.03	0
004	Fiber, Crude (%)	2073	13.41	0.1100	13.41	0.8105	0.2308	114	0.00	004.00	0
004	Fiber, Crude (%)	0849	13.42	0.1570	13.41	0.8105	0.2308	114	0.01	004.07	0
004	Fiber, Crude (%)	0229	13.44	0.1200	13.41	0.8105	0.2308	114	0.04	004.07	0
004	Fiber, Crude (%)	0100	13.46	0.3500	13.41	0.8105	0.2308	114	0.06	004.07	0
004	Fiber, Crude (%)	0003	13.53	0.5300	13.41	0.8105	0.2308	114	0.15	004.07	0
004	Fiber, Crude (%)	0893	13.53	0.0000	13.41	0.8105	0.2308	114	0.15	004.06	0
004	Fiber, Crude (%)	0164	13.55	0.3000	13.41	0.8105	0.2308	114	0.18	004.00	0
004	Fiber, Crude (%)	0848	13.55	0.0000	13.41	0.8105	0.2308	114	0.18	004.07	0
004	Fiber, Crude (%)	0918	13.56	0.2200	13.41	0.8105	0.2308	114	0.19	004.06	0
004	Fiber, Crude (%)	0309	13.58	0.0889	13.41	0.8105	0.2308	114	0.21	004.00	0
004	Fiber, Crude (%)	0878	13.66	0.0100	13.41	0.8105	0.2308	114	0.31	004.06	0
004	Fiber, Crude (%)	0870	13.67	0.3400	13.41	0.8105	0.2308	114	0.32	004.07	0
004	Fiber, Crude (%)	0885	13.69	0.1603	13.41	0.8105	0.2308	114	0.35	004.06	0
004	Fiber, Crude (%)	2152	13.72	0.0200	13.41	0.8105	0.2308	114	0.39	004.06	0
004	Fiber, Crude (%)	0074	13.74	0.1000	13.41	0.8105	0.2308	114	0.41	004.07	0
004	Fiber, Crude (%)	0674	13.75	0.0000	13.41	0.8105	0.2308	114	0.42	004.06	0
004	Fiber, Crude (%)	0689	13.75	0.1000	13.41	0.8105	0.2308	114	0.42	004.06	0
004	Fiber, Crude (%)	2192	13.76	0.1000	13.41	0.8105	0.2308	114	0.44	004.03	0
004	Fiber, Crude (%)	2259	13.77	0.1030	13.41	0.8105	0.2308	114	0.45	004.07	0
004	Fiber, Crude (%)	2006	13.85	0.0600	13.41	0.8105	0.2308	114	0.55	004.06	0
004	Fiber, Crude (%)	0675	13.90	0.0700	13.41	0.8105	0.2308	114	0.60	004.07	0
004	Fiber, Crude (%)	2053	13.93	0.0500	13.41	0.8105	0.2308	114	0.64	004.07	0
004	Fiber, Crude (%)	0720	13.95	0.0400	13.41	0.8105	0.2308	114	0.67	004.06	0
004	Fiber, Crude (%)	2193	13.95	0.0400	13.41	0.8105	0.2308	114	0.67	004.07	0
004	Fiber, Crude (%)	0563	13.97	0.3186	13.41	0.8105	0.2308	114	0.70	004.07	0
004	Fiber, Crude (%)	0171	13.99	0.5200	13.41	0.8105	0.2308	114	0.72	004.00	0
004	Fiber, Crude (%)	0226	14.01	0.1900	13.41	0.8105	0.2308	114	0.74	004.00	0
004	Fiber, Crude (%)	0034	14.01	0.1100	13.41	0.8105	0.2308	114	0.74	004.07	0
004	Fiber, Crude (%)	0676	14.05	0.0200	13.41	0.8105	0.2308	114	0.79	004.06	0
004	Fiber, Crude (%)	2076	14.07	0.4265	13.41	0.8105	0.2308	114	0.82	004.00	0
004	Fiber, Crude (%)	0505	14.26	0.1400	13.41	0.8105	0.2308	114	1.05	004.07	0
004	Fiber, Crude (%)	0675	14.31	0.0600	13.41	0.8105	0.2308	114	1.11	004.06	0
004	Fiber, Crude (%)	2109	14.32	0.3600	13.41	0.8105	0.2308	114	1.13	004.07	0
004	Fiber, Crude (%)	2196	14.35	0.0000	13.41	0.8105	0.2308	114	1.16	004.07	0
004	Fiber, Crude (%)	0066	14.36	0.1300	13.41	0.8105	0.2308	114	1.17	004.07	0
004	Fiber, Crude (%)	0948	14.37	0.2600	13.41	0.8105	0.2308	114	1.19	004.07	0
004	Fiber, Crude (%)	0504	14.39	0.4300	13.41	0.8105	0.2308	114	1.21	004.00	0
004	Fiber, Crude (%)	0948	14.41	0.1800	13.41	0.8105	0.2308	114	1.24	004.06	0
004	Fiber, Crude (%)	2264	14.55	0.3154	13.41	0.8105	0.2308	114	1.42	004.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
004	Fiber, Crude (%)	0693	14.72	0.4020	13.41	0.8105	0.2308	114	1.62	004.07	0
004	Fiber, Crude (%)	0407	14.75	0.8316	13.41	0.8105	0.2308	114	1.65	004.07	0
004	Fiber, Crude (%)	2246	15.11	0.3700	13.41	0.8105	0.2308	114	2.10	004.03	0
004	Fiber, Crude (%)	0876	15.14	0.7000	13.41	0.8105	0.2308	114	2.14	004.07	0
004	Fiber, Crude (%)	0656	15.60	0.6300	13.41	0.8105	0.2308	114	2.70	004.07	0
004	Fiber, Crude (%)	0529	15.77	0.4400	13.41	0.8105	0.2308	114	2.92	004.07	0
004	Fiber, Crude (%)	0598	15.98	1.120	13.41	0.8105	0.2308	114	3.17	004.07	0
004	Fiber, Crude (%)	0045	16.20	0.6000	13.41	0.8105	0.2308	114	3.45	004.07	0
004	Fiber, Crude (%)	0294	17.05	0.7000	13.41	0.8105	0.2308	114	4.49	004.07	0
004	Fiber, Crude (%)	0626	17.28	0.1300	13.41	0.8105	0.2308	114	4.77	004.07	0
004	Fiber, Crude (%)	0019	17.42	0.0100	13.41	0.8105	0.2308	114	4.95	004.07	0
004	Fiber, Crude (%)	0536	20.69	0.1100	13.41	0.8105	0.2308	114	8.98	004.07	0
004	Fiber, Crude (%)	0553	12.69	1.450	13.41	0.8105	0.2308	114	-0.89	004.11	1
004	Fiber, Crude (%)	0511	13.60	2.400	13.41	0.8105	0.2308	114	0.24	004.07	1
005	Ash (%)	0297	6.235	0.0300	10.44	0.3783	0.1009	141	-11.12	005.11	0
005	Ash (%)	2128	7.103	0.3249	10.44	0.3783	0.1009	141	-8.82	005.11	0
005	Ash (%)	0720	7.125	0.0300	10.44	0.3783	0.1009	141	-8.76	005.11	0
005	Ash (%)	0297	7.620	0.2200	10.44	0.3783	0.1009	141	-7.46	005.00	0
005	Ash (%)	0852	7.655	0.0700	10.44	0.3783	0.1009	141	-7.36	005.11	0
005	Ash (%)	2081	9.190	0.2200	10.44	0.3783	0.1009	141	-3.31	005.99	0
005	Ash (%)	2184	9.255	0.0300	10.44	0.3783	0.1009	141	-3.13	005.00	0
005	Ash (%)	2033	9.750	0.1000	10.44	0.3783	0.1009	141	-1.83	005.00	0
005	Ash (%)	0229	9.810	0.0800	10.44	0.3783	0.1009	141	-1.67	005.00	0
005	Ash (%)	0169	9.820	0.0400	10.44	0.3783	0.1009	141	-1.64	005.00	0
005	Ash (%)	0511	9.850	0.2400	10.44	0.3783	0.1009	141	-1.56	005.00	0
005	Ash (%)	2073	9.875	0.0500	10.44	0.3783	0.1009	141	-1.50	005.00	0
005	Ash (%)	0049	9.885	0.1100	10.44	0.3783	0.1009	141	-1.47	005.00	0
005	Ash (%)	0539	9.895	0.0300	10.44	0.3783	0.1009	141	-1.44	005.00	0
005	Ash (%)	0357	9.900	0.0000	10.44	0.3783	0.1009	141	-1.43	005.00	0
005	Ash (%)	0726	9.925	0.0100	10.44	0.3783	0.1009	141	-1.36	005.05	0
005	Ash (%)	0278	9.940	0.0400	10.44	0.3783	0.1009	141	-1.32	005.00	0
005	Ash (%)	0148	9.975	0.0300	10.44	0.3783	0.1009	141	-1.23	005.00	0
005	Ash (%)	0019	10.00	0.4000	10.44	0.3783	0.1009	141	-1.17	005.00	0
005	Ash (%)	0853	10.00	0.4000	10.44	0.3783	0.1009	141	-1.17	005.00	0
005	Ash (%)	0622	10.00	0.0428	10.44	0.3783	0.1009	141	-1.16	005.00	0
005	Ash (%)	0650	10.01	0.1500	10.44	0.3783	0.1009	141	-1.15	005.00	0
005	Ash (%)	0164	10.01	0.0900	10.44	0.3783	0.1009	141	-1.15	005.00	0
005	Ash (%)	0541	10.02	0.0500	10.44	0.3783	0.1009	141	-1.13	005.00	0
005	Ash (%)	0425	10.02	0.0400	10.44	0.3783	0.1009	141	-1.11	005.00	0
005	Ash (%)	0042	10.07	0.2000	10.44	0.3783	0.1009	141	-0.98	005.00	0
005	Ash (%)	0598	10.07	0.0200	10.44	0.3783	0.1009	141	-0.98	005.00	0
005	Ash (%)	0682	10.07	0.0000	10.44	0.3783	0.1009	141	-0.98	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0651	10.08	0.0050	10.44	0.3783	0.1009	141	-0.97	005.00	0
005	Ash (%)	2022	10.10	0.0100	10.44	0.3783	0.1009	141	-0.91	005.00	0
005	Ash (%)	0038	10.10	0.0800	10.44	0.3783	0.1009	141	-0.90	005.05	0
005	Ash (%)	0574	10.11	0.0700	10.44	0.3783	0.1009	141	-0.89	005.00	0
005	Ash (%)	0004	10.13	0.0100	10.44	0.3783	0.1009	141	-0.84	005.00	0
005	Ash (%)	0874	10.14	0.4400	10.44	0.3783	0.1009	141	-0.80	005.00	0
005	Ash (%)	0003	10.15	0.1000	10.44	0.3783	0.1009	141	-0.77	005.05	0
005	Ash (%)	0353	10.16	0.1700	10.44	0.3783	0.1009	141	-0.76	005.00	0
005	Ash (%)	0848	10.17	0.0000	10.44	0.3783	0.1009	141	-0.72	005.00	0
005	Ash (%)	0876	10.22	0.0950	10.44	0.3783	0.1009	141	-0.59	005.00	0
005	Ash (%)	2053	10.23	0.0100	10.44	0.3783	0.1009	141	-0.57	005.00	0
005	Ash (%)	0536	10.24	0.0000	10.44	0.3783	0.1009	141	-0.53	005.99	0
005	Ash (%)	0027	10.25	0.0600	10.44	0.3783	0.1009	141	-0.50	005.00	0
005	Ash (%)	0589	10.26	0.1000	10.44	0.3783	0.1009	141	-0.48	005.00	0
005	Ash (%)	0693	10.27	0.0110	10.44	0.3783	0.1009	141	-0.46	005.00	0
005	Ash (%)	0142	10.28	0.0600	10.44	0.3783	0.1009	141	-0.43	005.00	0
005	Ash (%)	2188	10.29	0.0500	10.44	0.3783	0.1009	141	-0.41	005.00	0
005	Ash (%)	2058	10.30	0.1300	10.44	0.3783	0.1009	141	-0.39	005.00	0
005	Ash (%)	0034	10.31	0.4610	10.44	0.3783	0.1009	141	-0.36	005.00	0
005	Ash (%)	0918	10.32	0.1900	10.44	0.3783	0.1009	141	-0.33	005.00	0
005	Ash (%)	0066	10.32	0.0600	10.44	0.3783	0.1009	141	-0.32	005.05	0
005	Ash (%)	0520	10.33	0.0200	10.44	0.3783	0.1009	141	-0.29	005.00	0
005	Ash (%)	0956	10.35	0.3000	10.44	0.3783	0.1009	141	-0.24	005.00	0
005	Ash (%)	2199	10.35	0.1000	10.44	0.3783	0.1009	141	-0.24	005.00	0
005	Ash (%)	0675	10.36	0.2600	10.44	0.3783	0.1009	141	-0.21	005.00	0
005	Ash (%)	0893	10.36	0.0000	10.44	0.3783	0.1009	141	-0.21	005.05	0
005	Ash (%)	0358	10.38	0.0400	10.44	0.3783	0.1009	141	-0.16	005.00	0
005	Ash (%)	0618	10.39	0.0978	10.44	0.3783	0.1009	141	-0.14	005.00	0
005	Ash (%)	0948	10.40	0.0500	10.44	0.3783	0.1009	141	-0.12	005.00	0
005	Ash (%)	0045	10.40	0.2000	10.44	0.3783	0.1009	141	-0.11	005.00	0
005	Ash (%)	0100	10.41	0.0800	10.44	0.3783	0.1009	141	-0.08	005.00	0
005	Ash (%)	0870	10.41	0.0939	10.44	0.3783	0.1009	141	-0.08	005.00	0
005	Ash (%)	0646	10.44	0.0500	10.44	0.3783	0.1009	141	-0.02	005.00	0
005	Ash (%)	2242	10.44	0.3900	10.44	0.3783	0.1009	141	-0.02	005.00	0
005	Ash (%)	0047	10.45	0.0400	10.44	0.3783	0.1009	141	0.02	005.00	0
005	Ash (%)	0298	10.45	0.0200	10.44	0.3783	0.1009	141	0.02	005.00	0
005	Ash (%)	0674	10.45	0.0000	10.44	0.3783	0.1009	141	0.02	005.00	0
005	Ash (%)	0619	10.45	0.1000	10.44	0.3783	0.1009	141	0.02	005.05	0
005	Ash (%)	0035	10.47	0.0500	10.44	0.3783	0.1009	141	0.06	005.00	0
005	Ash (%)	0083	10.48	0.0200	10.44	0.3783	0.1009	141	0.10	005.00	0
005	Ash (%)	2192	10.48	0.0800	10.44	0.3783	0.1009	141	0.10	005.05	0
005	Ash (%)	0953	10.49	0.4500	10.44	0.3783	0.1009	141	0.12	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0001	10.49	0.0227	10.44	0.3783	0.1009	141	0.13	005.00	0
005	Ash (%)	0033	10.50	0.0000	10.44	0.3783	0.1009	141	0.16	005.00	0
005	Ash (%)	2200	10.50	0.2000	10.44	0.3783	0.1009	141	0.16	005.00	0
005	Ash (%)	0638	10.50	0.2000	10.44	0.3783	0.1009	141	0.16	005.05	0
005	Ash (%)	0144	10.51	0.1700	10.44	0.3783	0.1009	141	0.17	005.00	0
005	Ash (%)	0108	10.51	0.1800	10.44	0.3783	0.1009	141	0.18	005.00	0
005	Ash (%)	0644	10.51	0.0000	10.44	0.3783	0.1009	141	0.18	005.02	0
005	Ash (%)	0171	10.52	0.0900	10.44	0.3783	0.1009	141	0.20	005.00	0
005	Ash (%)	0098	10.53	0.0500	10.44	0.3783	0.1009	141	0.22	005.00	0
005	Ash (%)	0123	10.53	0.1200	10.44	0.3783	0.1009	141	0.24	005.00	0
005	Ash (%)	2196	10.54	0.0000	10.44	0.3783	0.1009	141	0.26	005.00	0
005	Ash (%)	0510	10.56	0.0100	10.44	0.3783	0.1009	141	0.30	005.00	0
005	Ash (%)	2150	10.56	0.2300	10.44	0.3783	0.1009	141	0.30	005.00	0
005	Ash (%)	2258	10.57	0.2100	10.44	0.3783	0.1009	141	0.33	005.05	0
005	Ash (%)	2006	10.58	0.0500	10.44	0.3783	0.1009	141	0.35	005.00	0
005	Ash (%)	0202	10.59	0.0350	10.44	0.3783	0.1009	141	0.38	005.99	0
005	Ash (%)	0263	10.59	0.0343	10.44	0.3783	0.1009	141	0.39	005.05	0
005	Ash (%)	0505	10.60	0.0500	10.44	0.3783	0.1009	141	0.41	005.00	0
005	Ash (%)	0294	10.60	0.0000	10.44	0.3783	0.1009	141	0.42	005.05	0
005	Ash (%)	0407	10.61	0.0635	10.44	0.3783	0.1009	141	0.44	005.00	0
005	Ash (%)	0187	10.61	0.0200	10.44	0.3783	0.1009	141	0.45	005.00	0
005	Ash (%)	0889	10.61	0.2200	10.44	0.3783	0.1009	141	0.45	005.05	0
005	Ash (%)	0643	10.62	0.0900	10.44	0.3783	0.1009	141	0.46	005.00	0
005	Ash (%)	0968	10.62	0.0050	10.44	0.3783	0.1009	141	0.46	005.00	0
005	Ash (%)	2246	10.62	0.0400	10.44	0.3783	0.1009	141	0.47	005.00	0
005	Ash (%)	2144	10.63	0.0100	10.44	0.3783	0.1009	141	0.49	005.05	0
005	Ash (%)	0226	10.64	0.0900	10.44	0.3783	0.1009	141	0.51	005.00	0
005	Ash (%)	0366	10.65	0.1000	10.44	0.3783	0.1009	141	0.55	005.00	0
005	Ash (%)	2076	10.67	0.0259	10.44	0.3783	0.1009	141	0.60	005.00	0
005	Ash (%)	0563	10.67	0.0032	10.44	0.3783	0.1009	141	0.60	005.00	0
005	Ash (%)	0504	10.68	0.2000	10.44	0.3783	0.1009	141	0.63	005.00	0
005	Ash (%)	0689	10.68	0.0600	10.44	0.3783	0.1009	141	0.63	005.05	0
005	Ash (%)	0939	10.69	0.2300	10.44	0.3783	0.1009	141	0.64	005.05	0
005	Ash (%)	2062	10.69	0.0647	10.44	0.3783	0.1009	141	0.65	005.05	0
005	Ash (%)	0695	10.70	0.0300	10.44	0.3783	0.1009	141	0.67	005.05	0
005	Ash (%)	0652	10.70	0.0000	10.44	0.3783	0.1009	141	0.68	005.99	0
005	Ash (%)	0849	10.71	0.0150	10.44	0.3783	0.1009	141	0.70	005.05	0
005	Ash (%)	0683	10.72	0.4300	10.44	0.3783	0.1009	141	0.72	005.05	0
005	Ash (%)	2109	10.72	0.1000	10.44	0.3783	0.1009	141	0.74	005.05	0
005	Ash (%)	2146	10.72	0.0600	10.44	0.3783	0.1009	141	0.74	005.05	0
005	Ash (%)	0868	10.73	0.0300	10.44	0.3783	0.1009	141	0.75	005.05	0
005	Ash (%)	0885	10.73	0.0386	10.44	0.3783	0.1009	141	0.77	005.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0139	10.74	0.0300	10.44	0.3783	0.1009	141	0.78	005.00	0
005	Ash (%)	0656	10.75	0.2100	10.44	0.3783	0.1009	141	0.80	005.00	0
005	Ash (%)	0265	10.75	0.1000	10.44	0.3783	0.1009	141	0.82	005.00	0
005	Ash (%)	2190	10.75	0.3000	10.44	0.3783	0.1009	141	0.82	005.05	0
005	Ash (%)	2018	10.76	0.0500	10.44	0.3783	0.1009	141	0.83	005.05	0
005	Ash (%)	0062	10.76	0.3920	10.44	0.3783	0.1009	141	0.85	005.00	0
005	Ash (%)	0242	10.78	0.0700	10.44	0.3783	0.1009	141	0.88	005.00	0
005	Ash (%)	0590	10.79	0.0100	10.44	0.3783	0.1009	141	0.91	005.05	0
005	Ash (%)	2193	10.79	0.0000	10.44	0.3783	0.1009	141	0.92	005.00	0
005	Ash (%)	0897	10.79	0.0200	10.44	0.3783	0.1009	141	0.92	005.05	0
005	Ash (%)	0596	10.81	0.0200	10.44	0.3783	0.1009	141	0.98	005.05	0
005	Ash (%)	0676	10.82	0.2300	10.44	0.3783	0.1009	141	0.99	005.99	0
005	Ash (%)	2264	10.84	0.0286	10.44	0.3783	0.1009	141	1.05	005.05	0
005	Ash (%)	0878	10.85	0.1300	10.44	0.3783	0.1009	141	1.07	005.05	0
005	Ash (%)	0884	10.85	0.1000	10.44	0.3783	0.1009	141	1.08	005.99	0
005	Ash (%)	2152	10.86	0.0200	10.44	0.3783	0.1009	141	1.11	005.05	0
005	Ash (%)	2181	10.90	0.0770	10.44	0.3783	0.1009	141	1.22	005.00	0
005	Ash (%)	0581	10.91	0.1000	10.44	0.3783	0.1009	141	1.24	005.00	0
005	Ash (%)	2259	10.92	0.0700	10.44	0.3783	0.1009	141	1.27	005.00	0
005	Ash (%)	0309	10.94	0.0900	10.44	0.3783	0.1009	141	1.31	005.00	0
005	Ash (%)	2126	10.94	0.0200	10.44	0.3783	0.1009	141	1.32	005.05	0
005	Ash (%)	0345	10.97	0.0700	10.44	0.3783	0.1009	141	1.39	005.00	0
005	Ash (%)	0852	10.98	0.1600	10.44	0.3783	0.1009	141	1.42	005.00	0
005	Ash (%)	0610	10.98	0.0600	10.44	0.3783	0.1009	141	1.42	005.05	0
005	Ash (%)	2246	10.99	0.0400	10.44	0.3783	0.1009	141	1.45	005.05	0
005	Ash (%)	0190	11.01	0.0200	10.44	0.3783	0.1009	141	1.50	005.05	0
005	Ash (%)	0910	11.05	0.3000	10.44	0.3783	0.1009	141	1.61	005.99	0
005	Ash (%)	2089	11.10	0.1200	10.44	0.3783	0.1009	141	1.74	005.00	0
005	Ash (%)	2129	11.61	0.2100	10.44	0.3783	0.1009	141	3.08	005.05	0
005	Ash (%)	0553	10.26	0.6200	10.44	0.3783	0.1009	141	-0.48	005.00	1
005	Ash (%)	0720	< 0.01		10.44	0.3783	0.1009	141		005.05	5
006	Total Sugars (%)	0096	1.852	0.2286	3.571	1.193	0.5271	5		006.99	0
006	Total Sugars (%)	0910	2.850	1.360	3.571	1.193	0.5271	5		006.00	0
006	Total Sugars (%)	0407	3.997	0.3570	3.571	1.193	0.5271	5		006.00	0
006	Total Sugars (%)	0956	4.500	0.6000	3.571	1.193	0.5271	5		006.99	0
006	Total Sugars (%)	0720	4.655	0.0900	3.571	1.193	0.5271	5		006.99	0
008	Fiber, Acid Detergent (%)	0038	14.29	1.080	18.00	1.252	0.3163	61	-2.97	008.02	0
008	Fiber, Acid Detergent (%)	0956	15.70	0.0000	18.00	1.252	0.3163	61	-1.84	008.08	0
008	Fiber, Acid Detergent (%)	0720	15.76	0.0700	18.00	1.252	0.3163	61	-1.80	008.99	0
008	Fiber, Acid Detergent (%)	0676	15.84	0.1900	18.00	1.252	0.3163	61	-1.73	008.02	0
008	Fiber, Acid Detergent (%)	0619	15.90	0.2000	18.00	1.252	0.3163	61	-1.68	008.08	0
008	Fiber, Acid Detergent (%)	0910	15.94	0.1000	18.00	1.252	0.3163	61	-1.65	008.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
008	Fiber, Acid Detergent (%)	0590	16.14	0.2300	18.00	1.252	0.3163	61	-1.49	008.02	0
008	Fiber, Acid Detergent (%)	0848	16.36	0.0000	18.00	1.252	0.3163	61	-1.31	008.08	0
008	Fiber, Acid Detergent (%)	0049	16.43	0.0500	18.00	1.252	0.3163	61	-1.26	008.08	0
008	Fiber, Acid Detergent (%)	0878	16.68	0.3500	18.00	1.252	0.3163	61	-1.06	008.08	0
008	Fiber, Acid Detergent (%)	0035	16.75	0.0700	18.00	1.252	0.3163	61	-1.00	008.08	0
008	Fiber, Acid Detergent (%)	0004	16.92	0.8600	18.00	1.252	0.3163	61	-0.87	008.08	0
008	Fiber, Acid Detergent (%)	0610	17.08	0.6500	18.00	1.252	0.3163	61	-0.74	008.99	0
008	Fiber, Acid Detergent (%)	0948	17.12	0.3700	18.00	1.252	0.3163	61	-0.71	008.02	0
008	Fiber, Acid Detergent (%)	0968	17.12	0.2770	18.00	1.252	0.3163	61	-0.71	008.08	0
008	Fiber, Acid Detergent (%)	0278	17.20	0.4000	18.00	1.252	0.3163	61	-0.64	008.08	0
008	Fiber, Acid Detergent (%)	0581	17.33	0.2000	18.00	1.252	0.3163	61	-0.54	008.08	0
008	Fiber, Acid Detergent (%)	0939	17.35	0.0000	18.00	1.252	0.3163	61	-0.52	008.08	0
008	Fiber, Acid Detergent (%)	0148	17.43	0.0900	18.00	1.252	0.3163	61	-0.46	008.08	0
008	Fiber, Acid Detergent (%)	2196	17.52	0.0000	18.00	1.252	0.3163	61	-0.39	008.08	0
008	Fiber, Acid Detergent (%)	2190	17.63	0.0800	18.00	1.252	0.3163	61	-0.30	008.08	0
008	Fiber, Acid Detergent (%)	0294	17.65	0.3000	18.00	1.252	0.3163	61	-0.28	008.08	0
008	Fiber, Acid Detergent (%)	0885	17.71	0.1065	18.00	1.252	0.3163	61	-0.23	008.02	0
008	Fiber, Acid Detergent (%)	0263	17.87	0.0096	18.00	1.252	0.3163	61	-0.11	008.08	0
008	Fiber, Acid Detergent (%)	0098	17.90	0.2000	18.00	1.252	0.3163	61	-0.08	008.02	0
008	Fiber, Acid Detergent (%)	0164	17.95	0.1000	18.00	1.252	0.3163	61	-0.04	008.08	0
008	Fiber, Acid Detergent (%)	0353	17.95	0.3400	18.00	1.252	0.3163	61	-0.04	008.02	0
008	Fiber, Acid Detergent (%)	0358	17.99	0.1700	18.00	1.252	0.3163	61	-0.01	008.08	0
008	Fiber, Acid Detergent (%)	0407	18.00	1.417	18.00	1.252	0.3163	61	0.00	008.08	0
008	Fiber, Acid Detergent (%)	0190	18.01	0.0200	18.00	1.252	0.3163	61	0.01	008.08	0
008	Fiber, Acid Detergent (%)	0001	18.06	0.6900	18.00	1.252	0.3163	61	0.04	008.08	0
008	Fiber, Acid Detergent (%)	0870	18.10	1.129	18.00	1.252	0.3163	61	0.08	008.08	0
008	Fiber, Acid Detergent (%)	0646	18.16	0.2900	18.00	1.252	0.3163	61	0.12	008.08	0
008	Fiber, Acid Detergent (%)	2246	18.19	0.0500	18.00	1.252	0.3163	61	0.15	008.02	0
008	Fiber, Acid Detergent (%)	0675	18.24	0.2300	18.00	1.252	0.3163	61	0.19	008.08	0
008	Fiber, Acid Detergent (%)	0309	18.24	0.3106	18.00	1.252	0.3163	61	0.19	008.02	0
008	Fiber, Acid Detergent (%)	2192	18.30	1.130	18.00	1.252	0.3163	61	0.23	008.02	0
008	Fiber, Acid Detergent (%)	0357	18.35	0.1000	18.00	1.252	0.3163	61	0.28	008.08	0
008	Fiber, Acid Detergent (%)	0345	18.38	0.0700	18.00	1.252	0.3163	61	0.30	008.02	0
008	Fiber, Acid Detergent (%)	0066	18.49	0.1500	18.00	1.252	0.3163	61	0.39	008.08	0
008	Fiber, Acid Detergent (%)	0675	18.60	0.0300	18.00	1.252	0.3163	61	0.47	008.02	0
008	Fiber, Acid Detergent (%)	2193	18.77	0.2000	18.00	1.252	0.3163	61	0.61	008.08	0
008	Fiber, Acid Detergent (%)	0689	18.80	0.2000	18.00	1.252	0.3163	61	0.64	008.02	0
008	Fiber, Acid Detergent (%)	0504	18.81	0.4900	18.00	1.252	0.3163	61	0.64	008.02	0
008	Fiber, Acid Detergent (%)	0187	18.82	0.0600	18.00	1.252	0.3163	61	0.65	008.02	0
008	Fiber, Acid Detergent (%)	0083	18.87	0.0700	18.00	1.252	0.3163	61	0.69	008.08	0
008	Fiber, Acid Detergent (%)	0226	18.87	0.0200	18.00	1.252	0.3163	61	0.69	008.02	0
008	Fiber, Acid Detergent (%)	0037	18.90	0.3500	18.00	1.252	0.3163	61	0.71	008.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
008	Fiber, Acid Detergent (%)	0510	18.95	0.1000	18.00	1.252	0.3163	61	0.76	008.08	0
008	Fiber, Acid Detergent (%)	0964	18.96	0.0720	18.00	1.252	0.3163	61	0.76	008.02	0
008	Fiber, Acid Detergent (%)	2129	19.04	1.760	18.00	1.252	0.3163	61	0.83	008.08	0
008	Fiber, Acid Detergent (%)	0148	19.09	0.2400	18.00	1.252	0.3163	61	0.87	008.02	0
008	Fiber, Acid Detergent (%)	0918	19.11	0.2420	18.00	1.252	0.3163	61	0.88	008.08	0
008	Fiber, Acid Detergent (%)	0596	19.22	0.3500	18.00	1.252	0.3163	61	0.97	008.08	0
008	Fiber, Acid Detergent (%)	0032	19.28	0.6600	18.00	1.252	0.3163	61	1.02	008.08	0
008	Fiber, Acid Detergent (%)	0045	19.70	0.4000	18.00	1.252	0.3163	61	1.36	008.08	0
008	Fiber, Acid Detergent (%)	0726	19.73	0.0900	18.00	1.252	0.3163	61	1.38	008.08	0
008	Fiber, Acid Detergent (%)	0693	19.77	0.2640	18.00	1.252	0.3163	61	1.41	008.08	0
008	Fiber, Acid Detergent (%)	0536	19.89	0.0400	18.00	1.252	0.3163	61	1.51	008.08	0
008	Fiber, Acid Detergent (%)	0265	19.90	0.8000	18.00	1.252	0.3163	61	1.52	008.05	0
008	Fiber, Acid Detergent (%)	0015	20.14	0.7800	18.00	1.252	0.3163	61	1.71	008.08	0
008	Fiber, Acid Detergent (%)	0511	17.00	3.000	18.00	1.252	0.3163	61	-0.80	008.99	1
008	Fiber, Acid Detergent (%)	0297	5.060	0.2400	18.00	1.252	0.3163	61	-10.34	008.08	2
009	Fiber, Neutral Detergent (%)	0619	30.70	0.0000	35.32	1.749	0.4536	61	-2.64	009.09	0
009	Fiber, Neutral Detergent (%)	0720	31.41	0.3700	35.32	1.749	0.4536	61	-2.24	009.99	0
009	Fiber, Neutral Detergent (%)	0263	31.78	0.0777	35.32	1.749	0.4536	61	-2.02	009.09	0
009	Fiber, Neutral Detergent (%)	0049	32.00	0.3100	35.32	1.749	0.4536	61	-1.90	009.09	0
009	Fiber, Neutral Detergent (%)	0918	32.56	1.813	35.32	1.749	0.4536	61	-1.58	009.09	0
009	Fiber, Neutral Detergent (%)	0910	32.80	0.8000	35.32	1.749	0.4536	61	-1.44	009.07	0
009	Fiber, Neutral Detergent (%)	0956	32.80	0.0000	35.32	1.749	0.4536	61	-1.44	009.09	0
009	Fiber, Neutral Detergent (%)	0407	33.51	0.2833	35.32	1.749	0.4536	61	-1.04	009.09	0
009	Fiber, Neutral Detergent (%)	0164	33.55	0.3000	35.32	1.749	0.4536	61	-1.01	009.09	0
009	Fiber, Neutral Detergent (%)	0598	33.60	0.5300	35.32	1.749	0.4536	61	-0.99	009.09	0
009	Fiber, Neutral Detergent (%)	0939	33.70	0.7800	35.32	1.749	0.4536	61	-0.93	009.09	0
009	Fiber, Neutral Detergent (%)	0278	33.85	0.1000	35.32	1.749	0.4536	61	-0.84	009.09	0
009	Fiber, Neutral Detergent (%)	0358	34.13	0.0900	35.32	1.749	0.4536	61	-0.68	009.09	0
009	Fiber, Neutral Detergent (%)	0510	34.20	1.000	35.32	1.749	0.4536	61	-0.64	009.09	0
009	Fiber, Neutral Detergent (%)	0870	34.30	0.4426	35.32	1.749	0.4536	61	-0.59	009.09	0
009	Fiber, Neutral Detergent (%)	0357	34.30	0.6000	35.32	1.749	0.4536	61	-0.58	009.09	0
009	Fiber, Neutral Detergent (%)	0878	34.32	0.3500	35.32	1.749	0.4536	61	-0.57	009.09	0
009	Fiber, Neutral Detergent (%)	0581	34.33	0.0200	35.32	1.749	0.4536	61	-0.57	009.09	0
009	Fiber, Neutral Detergent (%)	0294	34.40	0.2000	35.32	1.749	0.4536	61	-0.53	009.09	0
009	Fiber, Neutral Detergent (%)	2190	34.56	0.3200	35.32	1.749	0.4536	61	-0.43	009.09	0
009	Fiber, Neutral Detergent (%)	2129	34.67	0.7300	35.32	1.749	0.4536	61	-0.37	009.09	0
009	Fiber, Neutral Detergent (%)	0098	34.90	1.800	35.32	1.749	0.4536	61	-0.24	009.07	0
009	Fiber, Neutral Detergent (%)	0590	34.94	0.0400	35.32	1.749	0.4536	61	-0.22	009.07	0
009	Fiber, Neutral Detergent (%)	0848	34.96	0.0000	35.32	1.749	0.4536	61	-0.21	009.09	0
009	Fiber, Neutral Detergent (%)	0676	35.08	0.2400	35.32	1.749	0.4536	61	-0.14	009.07	0
009	Fiber, Neutral Detergent (%)	0187	35.12	0.1300	35.32	1.749	0.4536	61	-0.12	009.07	0
009	Fiber, Neutral Detergent (%)	0596	35.20	0.1300	35.32	1.749	0.4536	61	-0.07	009.09	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
009	Fiber, Neutral Detergent (%)	2196	35.21	0.0000	35.32	1.749	0.4536	61	-0.06	009.09	0
009	Fiber, Neutral Detergent (%)	0646	35.22	0.2200	35.32	1.749	0.4536	61	-0.06	009.09	0
009	Fiber, Neutral Detergent (%)	0001	35.25	0.6700	35.32	1.749	0.4536	61	-0.04	009.09	0
009	Fiber, Neutral Detergent (%)	0968	35.31	1.302	35.32	1.749	0.4536	61	-0.01	009.09	0
009	Fiber, Neutral Detergent (%)	0083	35.39	0.0200	35.32	1.749	0.4536	61	0.04	009.09	0
009	Fiber, Neutral Detergent (%)	0504	35.44	1.160	35.32	1.749	0.4536	61	0.07	009.04	0
009	Fiber, Neutral Detergent (%)	2193	35.54	0.4500	35.32	1.749	0.4536	61	0.12	009.09	0
009	Fiber, Neutral Detergent (%)	0265	35.55	0.1000	35.32	1.749	0.4536	61	0.13	009.09	0
009	Fiber, Neutral Detergent (%)	0015	35.61	0.1700	35.32	1.749	0.4536	61	0.16	009.09	0
009	Fiber, Neutral Detergent (%)	0226	35.63	0.1000	35.32	1.749	0.4536	61	0.18	009.07	0
009	Fiber, Neutral Detergent (%)	0948	35.67	0.3200	35.32	1.749	0.4536	61	0.20	009.09	0
009	Fiber, Neutral Detergent (%)	0037	35.69	0.2200	35.32	1.749	0.4536	61	0.21	009.09	0
009	Fiber, Neutral Detergent (%)	0693	35.81	0.7540	35.32	1.749	0.4536	61	0.28	009.09	0
009	Fiber, Neutral Detergent (%)	0032	35.87	0.4200	35.32	1.749	0.4536	61	0.31	009.09	0
009	Fiber, Neutral Detergent (%)	0675	35.90	0.3600	35.32	1.749	0.4536	61	0.33	009.09	0
009	Fiber, Neutral Detergent (%)	0353	36.06	0.1700	35.32	1.749	0.4536	61	0.42	009.07	0
009	Fiber, Neutral Detergent (%)	0345	36.07	0.2200	35.32	1.749	0.4536	61	0.43	009.07	0
009	Fiber, Neutral Detergent (%)	0610	36.15	0.5000	35.32	1.749	0.4536	61	0.47	009.07	0
009	Fiber, Neutral Detergent (%)	0638	36.20	0.2000	35.32	1.749	0.4536	61	0.50	009.09	0
009	Fiber, Neutral Detergent (%)	0885	36.48	0.2812	35.32	1.749	0.4536	61	0.66	009.07	0
009	Fiber, Neutral Detergent (%)	0066	36.52	0.5700	35.32	1.749	0.4536	61	0.68	009.09	0
009	Fiber, Neutral Detergent (%)	0675	36.52	0.1400	35.32	1.749	0.4536	61	0.69	009.07	0
009	Fiber, Neutral Detergent (%)	0148	36.57	0.4200	35.32	1.749	0.4536	61	0.71	009.09	0
009	Fiber, Neutral Detergent (%)	0964	36.74	0.7220	35.32	1.749	0.4536	61	0.81	009.07	0
009	Fiber, Neutral Detergent (%)	0045	36.95	1.900	35.32	1.749	0.4536	61	0.93	009.09	0
009	Fiber, Neutral Detergent (%)	2192	37.01	2.050	35.32	1.749	0.4536	61	0.96	009.07	0
009	Fiber, Neutral Detergent (%)	0190	37.93	1.350	35.32	1.749	0.4536	61	1.49	009.09	0
009	Fiber, Neutral Detergent (%)	0726	38.11	0.0100	35.32	1.749	0.4536	61	1.59	009.09	0
009	Fiber, Neutral Detergent (%)	0309	38.22	0.2260	35.32	1.749	0.4536	61	1.66	009.07	0
009	Fiber, Neutral Detergent (%)	0689	38.30	0.4000	35.32	1.749	0.4536	61	1.70	009.07	0
009	Fiber, Neutral Detergent (%)	0941	38.31	0.0400	35.32	1.749	0.4536	61	1.71	009.99	0
009	Fiber, Neutral Detergent (%)	2246	39.10	0.0300	35.32	1.749	0.4536	61	2.16	009.04	0
009	Fiber, Neutral Detergent (%)	0536	39.92	0.1200	35.32	1.749	0.4536	61	2.63	009.09	0
009	Fiber, Neutral Detergent (%)	2246	43.15	0.6000	35.32	1.749	0.4536	61	4.48	009.07	0
009	Fiber, Neutral Detergent (%)	0297	18.20	0.2000	35.32	1.749	0.4536	61	-9.79	009.09	2
010	Moisture (%)	0726	7.525	0.0300	8.495	0.6423	0.0665	18	-1.51	010.99	0
010	Moisture (%)	0941	7.715	0.0500	8.495	0.6423	0.0665	18	-1.21	010.99	0
010	Moisture (%)	0843	7.895	0.1500	8.495	0.6423	0.0665	18	-0.93	010.03	0
010	Moisture (%)	2178	8.130	0.1400	8.495	0.6423	0.0665	18	-0.57	010.99	0
010	Moisture (%)	0948	8.145	0.0100	8.495	0.6423	0.0665	18	-0.55	010.99	0
010	Moisture (%)	0164	8.260	0.0800	8.495	0.6423	0.0665	18	-0.37	010.03	0
010	Moisture (%)	0939	8.300	0.0400	8.495	0.6423	0.0665	18	-0.30	010.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
010	Moisture (%)	2043	8.350	0.0400	8.495	0.6423	0.0665	18	-0.23	010.99	0
010	Moisture (%)	0652	8.350	0.1000	8.495	0.6423	0.0665	18	-0.23	010.99	0
010	Moisture (%)	0884	8.350	0.1000	8.495	0.6423	0.0665	18	-0.23	010.99	0
010	Moisture (%)	2150	8.395	0.0300	8.495	0.6423	0.0665	18	-0.16	010.99	0
010	Moisture (%)	2089	8.615	0.0500	8.495	0.6423	0.0665	18	0.19	010.99	0
010	Moisture (%)	0190	8.770	0.0600	8.495	0.6423	0.0665	18	0.43	010.99	0
010	Moisture (%)	2076	8.770	0.0400	8.495	0.6423	0.0665	18	0.43	010.99	0
010	Moisture (%)	2128	8.993	0.1466	8.495	0.6423	0.0665	18	0.77	010.11	0
010	Moisture (%)	0720	9.425	0.0300	8.495	0.6423	0.0665	18	1.45	010.11	0
010	Moisture (%)	0405	10.02	0.0000	8.495	0.6423	0.0665	18	2.37	010.99	0
010	Moisture (%)	0852	11.25	0.1000	8.495	0.6423	0.0665	18	4.29	010.11	0
010	Moisture (%)	2129	7.900	0.5000	8.495	0.6423	0.0665	18	-0.93	010.99	1
011	Loss on Drying (%)	2031	5.633	0.0150	9.058	0.4394	0.1003	68	-7.80	011.01	0
011	Loss on Drying (%)	0618	6.555	0.1300	9.058	0.4394	0.1003	68	-5.70	011.01	0
011	Loss on Drying (%)	2181	8.120	0.0200	9.058	0.4394	0.1003	68	-2.13	011.01	0
011	Loss on Drying (%)	2200	8.150	0.1000	9.058	0.4394	0.1003	68	-2.07	011.01	0
011	Loss on Drying (%)	2199	8.200	0.2000	9.058	0.4394	0.1003	68	-1.95	011.01	0
011	Loss on Drying (%)	0650	8.255	0.0700	9.058	0.4394	0.1003	68	-1.83	011.01	0
011	Loss on Drying (%)	0646	8.330	0.1400	9.058	0.4394	0.1003	68	-1.66	011.01	0
011	Loss on Drying (%)	0563	8.355	0.1390	9.058	0.4394	0.1003	68	-1.60	011.01	0
011	Loss on Drying (%)	0874	8.435	0.0300	9.058	0.4394	0.1003	68	-1.42	011.01	0
011	Loss on Drying (%)	0968	8.499	0.0050	9.058	0.4394	0.1003	68	-1.27	011.01	0
011	Loss on Drying (%)	0263	8.560	0.0800	9.058	0.4394	0.1003	68	-1.13	011.01	0
011	Loss on Drying (%)	0598	8.630	0.1400	9.058	0.4394	0.1003	68	-0.97	011.01	0
011	Loss on Drying (%)	0675	8.695	0.0500	9.058	0.4394	0.1003	68	-0.83	011.01	0
011	Loss on Drying (%)	2188	8.740	0.0600	9.058	0.4394	0.1003	68	-0.72	011.01	0
011	Loss on Drying (%)	0651	8.809	0.0640	9.058	0.4394	0.1003	68	-0.57	011.01	0
011	Loss on Drying (%)	2168	8.810	0.2200	9.058	0.4394	0.1003	68	-0.56	011.01	0
011	Loss on Drying (%)	0541	8.850	0.3600	9.058	0.4394	0.1003	68	-0.47	011.01	0
011	Loss on Drying (%)	0520	8.850	0.1000	9.058	0.4394	0.1003	68	-0.47	011.01	0
011	Loss on Drying (%)	0857	8.850	0.1000	9.058	0.4394	0.1003	68	-0.47	011.99	0
011	Loss on Drying (%)	0047	8.860	0.2000	9.058	0.4394	0.1003	68	-0.45	011.01	0
011	Loss on Drying (%)	0298	8.870	0.0200	9.058	0.4394	0.1003	68	-0.43	011.01	0
011	Loss on Drying (%)	0536	8.890	0.1800	9.058	0.4394	0.1003	68	-0.38	011.01	0
011	Loss on Drying (%)	2033	8.895	0.0300	9.058	0.4394	0.1003	68	-0.37	011.01	0
011	Loss on Drying (%)	0859	8.897	0.0990	9.058	0.4394	0.1003	68	-0.37	011.01	0
011	Loss on Drying (%)	0674	8.920	0.0000	9.058	0.4394	0.1003	68	-0.31	011.01	0
011	Loss on Drying (%)	0942	8.920	0.4200	9.058	0.4394	0.1003	68	-0.31	011.02	0
011	Loss on Drying (%)	0574	8.925	0.0300	9.058	0.4394	0.1003	68	-0.30	011.01	0
011	Loss on Drying (%)	0953	8.940	0.2200	9.058	0.4394	0.1003	68	-0.27	011.01	0
011	Loss on Drying (%)	2073	8.970	0.0400	9.058	0.4394	0.1003	68	-0.20	011.01	0
011	Loss on Drying (%)	0553	8.990	0.4400	9.058	0.4394	0.1003	68	-0.15	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
011	Loss on Drying (%)	0100	9.055	0.0500	9.058	0.4394	0.1003	68	-0.01	011.01	0
011	Loss on Drying (%)	2196	9.090	0.0000	9.058	0.4394	0.1003	68	0.07	011.01	0
011	Loss on Drying (%)	2246	9.090	0.0800	9.058	0.4394	0.1003	68	0.07	011.01	0
011	Loss on Drying (%)	0265	9.100	0.0000	9.058	0.4394	0.1003	68	0.10	011.01	0
011	Loss on Drying (%)	0510	9.100	0.2000	9.058	0.4394	0.1003	68	0.10	011.01	0
011	Loss on Drying (%)	0229	9.129	0.0110	9.058	0.4394	0.1003	68	0.16	011.01	0
011	Loss on Drying (%)	0062	9.134	0.0570	9.058	0.4394	0.1003	68	0.17	011.01	0
011	Loss on Drying (%)	0407	9.138	0.0949	9.058	0.4394	0.1003	68	0.18	011.01	0
011	Loss on Drying (%)	0226	9.150	0.1000	9.058	0.4394	0.1003	68	0.21	011.01	0
011	Loss on Drying (%)	0294	9.150	0.1000	9.058	0.4394	0.1003	68	0.21	011.01	0
011	Loss on Drying (%)	0539	9.150	0.0000	9.058	0.4394	0.1003	68	0.21	011.01	0
011	Loss on Drying (%)	2184	9.155	0.0500	9.058	0.4394	0.1003	68	0.22	011.01	0
011	Loss on Drying (%)	0511	9.170	0.1200	9.058	0.4394	0.1003	68	0.26	011.01	0
011	Loss on Drying (%)	2193	9.180	0.1600	9.058	0.4394	0.1003	68	0.28	011.01	0
011	Loss on Drying (%)	0682	9.200	0.0000	9.058	0.4394	0.1003	68	0.32	011.01	0
011	Loss on Drying (%)	0164	9.245	0.0500	9.058	0.4394	0.1003	68	0.43	011.01	0
011	Loss on Drying (%)	2006	9.255	0.0500	9.058	0.4394	0.1003	68	0.45	011.01	0
011	Loss on Drying (%)	0098	9.265	0.1500	9.058	0.4394	0.1003	68	0.47	011.01	0
011	Loss on Drying (%)	0425	9.270	0.0400	9.058	0.4394	0.1003	68	0.48	011.01	0
011	Loss on Drying (%)	0202	9.330	0.0510	9.058	0.4394	0.1003	68	0.62	011.01	0
011	Loss on Drying (%)	0589	9.340	0.1800	9.058	0.4394	0.1003	68	0.64	011.01	0
011	Loss on Drying (%)	2057	9.344	0.0036	9.058	0.4394	0.1003	68	0.65	011.01	0
011	Loss on Drying (%)	0870	9.345	0.0355	9.058	0.4394	0.1003	68	0.65	011.01	0
011	Loss on Drying (%)	0144	9.350	0.0400	9.058	0.4394	0.1003	68	0.67	011.01	0
011	Loss on Drying (%)	0171	9.400	0.0200	9.058	0.4394	0.1003	68	0.78	011.01	0
011	Loss on Drying (%)	0108	9.435	0.2100	9.058	0.4394	0.1003	68	0.86	011.01	0
011	Loss on Drying (%)	0242	9.450	0.0600	9.058	0.4394	0.1003	68	0.89	011.01	0
011	Loss on Drying (%)	2022	9.475	0.0300	9.058	0.4394	0.1003	68	0.95	011.01	0
011	Loss on Drying (%)	0205	9.542	0.0330	9.058	0.4394	0.1003	68	1.10	011.01	0
011	Loss on Drying (%)	0622	9.560	0.1756	9.058	0.4394	0.1003	68	1.14	011.01	0
011	Loss on Drying (%)	0358	9.585	0.0100	9.058	0.4394	0.1003	68	1.20	011.01	0
011	Loss on Drying (%)	0123	9.590	0.1600	9.058	0.4394	0.1003	68	1.21	011.01	0
011	Loss on Drying (%)	0066	9.620	0.2200	9.058	0.4394	0.1003	68	1.28	011.01	0
011	Loss on Drying (%)	0004	9.665	0.0300	9.058	0.4394	0.1003	68	1.38	011.01	0
011	Loss on Drying (%)	0309	9.730	0.1200	9.058	0.4394	0.1003	68	1.53	011.01	0
011	Loss on Drying (%)	0843	9.770	0.1800	9.058	0.4394	0.1003	68	1.62	011.01	0
011	Loss on Drying (%)	2058	9.860	0.1600	9.058	0.4394	0.1003	68	1.83	011.01	0
011	Loss on Drying (%)	0852	10.12	0.0900	9.058	0.4394	0.1003	68	2.41	011.99	0
011	Loss on Drying (%)	0643	8.200	0.8000	9.058	0.4394	0.1003	68	-1.95	011.01	1
012	Starch (%)	0676	14.25	0.5500	16.26	1.254	0.2457	38	-1.61	012.01	0
012	Starch (%)	0510	14.40	0.0000	16.26	1.254	0.2457	38	-1.49	012.04	0
012	Starch (%)	0956	14.80	0.2000	16.26	1.254	0.2457	38	-1.17	012.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
012	Starch (%)	0098	14.80	0.0000	16.26	1.254	0.2457	38	-1.17	012.03	0
012	Starch (%)	0407	14.83	0.0610	16.26	1.254	0.2457	38	-1.15	012.03	0
012	Starch (%)	0164	15.00	0.4000	16.26	1.254	0.2457	38	-1.01	012.01	0
012	Starch (%)	0870	15.09	0.2080	16.26	1.254	0.2457	38	-0.94	012.01	0
012	Starch (%)	0278	15.15	0.1000	16.26	1.254	0.2457	38	-0.89	012.04	0
012	Starch (%)	0148	15.27	0.9400	16.26	1.254	0.2457	38	-0.79	012.01	0
012	Starch (%)	0045	15.35	0.9000	16.26	1.254	0.2457	38	-0.73	012.01	0
012	Starch (%)	0353	15.45	0.1500	16.26	1.254	0.2457	38	-0.65	012.04	0
012	Starch (%)	0265	15.60	0.0000	16.26	1.254	0.2457	38	-0.53	012.01	0
012	Starch (%)	0939	15.85	0.0500	16.26	1.254	0.2457	38	-0.33	012.00	0
012	Starch (%)	0720	15.88	0.1600	16.26	1.254	0.2457	38	-0.30	012.11	0
012	Starch (%)	0889	16.08	0.6800	16.26	1.254	0.2457	38	-0.15	012.03	0
012	Starch (%)	0066	16.15	0.3000	16.26	1.254	0.2457	38	-0.09	012.03	0
012	Starch (%)	0638	16.20	0.0000	16.26	1.254	0.2457	38	-0.05	012.00	0
012	Starch (%)	0689	16.25	0.1000	16.26	1.254	0.2457	38	-0.01	012.00	0
012	Starch (%)	0885	16.26	0.0509	16.26	1.254	0.2457	38	0.00	012.01	0
012	Starch (%)	2006	16.30	0.4100	16.26	1.254	0.2457	38	0.03	012.00	0
012	Starch (%)	2193	16.38	0.3100	16.26	1.254	0.2457	38	0.09	012.00	0
012	Starch (%)	0693	16.43	0.2770	16.26	1.254	0.2457	38	0.14	012.01	0
012	Starch (%)	2242	16.45	0.5000	16.26	1.254	0.2457	38	0.15	012.01	0
012	Starch (%)	0918	16.49	1.223	16.26	1.254	0.2457	38	0.18	012.01	0
012	Starch (%)	0610	16.50	0.2000	16.26	1.254	0.2457	38	0.19	012.00	0
012	Starch (%)	0619	16.50	0.0000	16.26	1.254	0.2457	38	0.19	012.00	0
012	Starch (%)	2188	16.52	0.0200	16.26	1.254	0.2457	38	0.21	012.00	0
012	Starch (%)	2196	16.57	0.0000	16.26	1.254	0.2457	38	0.25	012.00	0
012	Starch (%)	0164	16.70	0.4000	16.26	1.254	0.2457	38	0.35	012.00	0
012	Starch (%)	0878	17.00	0.2000	16.26	1.254	0.2457	38	0.59	012.00	0
012	Starch (%)	0948	17.04	0.2600	16.26	1.254	0.2457	38	0.62	012.04	0
012	Starch (%)	0683	17.45	0.2200	16.26	1.254	0.2457	38	0.95	012.00	0
012	Starch (%)	2089	18.32	0.1400	16.26	1.254	0.2457	38	1.64	012.01	0
012	Starch (%)	2192	18.37	0.0100	16.26	1.254	0.2457	38	1.68	012.00	0
012	Starch (%)	2128	22.37	0.0860	16.26	1.254	0.2457	38	4.87	012.11	0
012	Starch (%)	0297	24.44	0.1500	16.26	1.254	0.2457	38	6.52	012.11	0
012	Starch (%)	0852	27.78	0.0600	16.26	1.254	0.2457	38	9.19	012.11	0
012	Starch (%)	0297	33.71	0.0200	16.26	1.254	0.2457	38	13.92	012.03	0
012	Starch (%)	0910	15.40	2.000	16.26	1.254	0.2457	38	-0.69	012.00	1
013	Fat, Acid Pretreat (%)	0618	2.937	0.0904	4.463	0.6717	0.1530	51	-2.27	013.08	0
013	Fat, Acid Pretreat (%)	0618	3.303	0.1067	4.463	0.6717	0.1530	51	-1.73	013.00	0
013	Fat, Acid Pretreat (%)	0720	3.575	0.4700	4.463	0.6717	0.1530	51	-1.32	013.10	0
013	Fat, Acid Pretreat (%)	0581	3.610	0.0800	4.463	0.6717	0.1530	51	-1.27	013.13	0
013	Fat, Acid Pretreat (%)	2073	3.635	0.0500	4.463	0.6717	0.1530	51	-1.23	013.00	0
013	Fat, Acid Pretreat (%)	0083	3.675	0.0900	4.463	0.6717	0.1530	51	-1.17	013.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
013	Fat, Acid Pretreat (%)	0036	3.678	0.1013	4.463	0.6717	0.1530	51	-1.17	013.10	0
013	Fat, Acid Pretreat (%)	2128	3.683	0.0402	4.463	0.6717	0.1530	51	-1.16	013.12	0
013	Fat, Acid Pretreat (%)	0689	3.750	0.1000	4.463	0.6717	0.1530	51	-1.06	013.00	0
013	Fat, Acid Pretreat (%)	0504	3.895	0.1100	4.463	0.6717	0.1530	51	-0.85	013.00	0
013	Fat, Acid Pretreat (%)	2146	3.940	0.1400	4.463	0.6717	0.1530	51	-0.78	013.00	0
013	Fat, Acid Pretreat (%)	0229	3.955	0.0900	4.463	0.6717	0.1530	51	-0.76	013.00	0
013	Fat, Acid Pretreat (%)	0242	3.980	0.5600	4.463	0.6717	0.1530	51	-0.72	013.00	0
013	Fat, Acid Pretreat (%)	2196	4.000	0.0000	4.463	0.6717	0.1530	51	-0.69	013.10	0
013	Fat, Acid Pretreat (%)	0843	4.085	0.1500	4.463	0.6717	0.1530	51	-0.56	013.13	0
013	Fat, Acid Pretreat (%)	2058	4.160	0.0600	4.463	0.6717	0.1530	51	-0.45	013.02	0
013	Fat, Acid Pretreat (%)	0610	4.160	0.0200	4.463	0.6717	0.1530	51	-0.45	013.10	0
013	Fat, Acid Pretreat (%)	0098	4.175	0.0500	4.463	0.6717	0.1530	51	-0.43	013.02	0
013	Fat, Acid Pretreat (%)	0948	4.220	0.1800	4.463	0.6717	0.1530	51	-0.36	013.00	0
013	Fat, Acid Pretreat (%)	0910	4.245	0.3500	4.463	0.6717	0.1530	51	-0.32	013.00	0
013	Fat, Acid Pretreat (%)	0650	4.260	0.6000	4.463	0.6717	0.1530	51	-0.30	013.02	0
013	Fat, Acid Pretreat (%)	2057	4.324	0.1474	4.463	0.6717	0.1530	51	-0.21	013.13	0
013	Fat, Acid Pretreat (%)	2006	4.390	0.1200	4.463	0.6717	0.1530	51	-0.11	013.10	0
013	Fat, Acid Pretreat (%)	0853	4.400	0.2000	4.463	0.6717	0.1530	51	-0.09	013.02	0
013	Fat, Acid Pretreat (%)	0870	4.410	0.1800	4.463	0.6717	0.1530	51	-0.08	013.02	0
013	Fat, Acid Pretreat (%)	0353	4.440	0.0800	4.463	0.6717	0.1530	51	-0.03	013.10	0
013	Fat, Acid Pretreat (%)	0884	4.450	0.1000	4.463	0.6717	0.1530	51	-0.02	013.00	0
013	Fat, Acid Pretreat (%)	0100	4.460	0.2000	4.463	0.6717	0.1530	51	0.00	013.02	0
013	Fat, Acid Pretreat (%)	0297	4.470	0.0800	4.463	0.6717	0.1530	51	0.01	013.00	0
013	Fat, Acid Pretreat (%)	0652	4.500	0.0000	4.463	0.6717	0.1530	51	0.05	013.00	0
013	Fat, Acid Pretreat (%)	0309	4.596	0.0353	4.463	0.6717	0.1530	51	0.20	013.00	0
013	Fat, Acid Pretreat (%)	0541	4.645	0.2300	4.463	0.6717	0.1530	51	0.27	013.00	0
013	Fat, Acid Pretreat (%)	2264	4.662	0.1596	4.463	0.6717	0.1530	51	0.30	013.02	0
013	Fat, Acid Pretreat (%)	2150	4.745	0.1300	4.463	0.6717	0.1530	51	0.42	013.00	0
013	Fat, Acid Pretreat (%)	0148	4.765	0.0700	4.463	0.6717	0.1530	51	0.45	013.02	0
013	Fat, Acid Pretreat (%)	0675	4.815	0.3300	4.463	0.6717	0.1530	51	0.52	013.02	0
013	Fat, Acid Pretreat (%)	0682	4.840	0.0000	4.463	0.6717	0.1530	51	0.56	013.02	0
013	Fat, Acid Pretreat (%)	0171	4.865	0.3300	4.463	0.6717	0.1530	51	0.60	013.02	0
013	Fat, Acid Pretreat (%)	0139	5.060	0.1000	4.463	0.6717	0.1530	51	0.89	013.00	0
013	Fat, Acid Pretreat (%)	0164	5.100	0.1000	4.463	0.6717	0.1530	51	0.95	013.02	0
013	Fat, Acid Pretreat (%)	0202	5.107	0.1220	4.463	0.6717	0.1530	51	0.96	013.00	0
013	Fat, Acid Pretreat (%)	2076	5.123	0.2518	4.463	0.6717	0.1530	51	0.98	013.00	0
013	Fat, Acid Pretreat (%)	0553	5.130	0.0600	4.463	0.6717	0.1530	51	0.99	013.02	0
013	Fat, Acid Pretreat (%)	0676	5.165	0.0500	4.463	0.6717	0.1530	51	1.04	013.02	0
013	Fat, Acid Pretreat (%)	2259	5.189	0.2740	4.463	0.6717	0.1530	51	1.08	013.02	0
013	Fat, Acid Pretreat (%)	0407	5.222	0.2055	4.463	0.6717	0.1530	51	1.13	013.02	0
013	Fat, Acid Pretreat (%)	2033	5.275	0.1500	4.463	0.6717	0.1530	51	1.21	013.02	0
013	Fat, Acid Pretreat (%)	0425	5.570	0.0200	4.463	0.6717	0.1530	51	1.65	013.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
013	Fat, Acid Pretreat (%)	0265	6.055	0.0100	4.463	0.6717	0.1530	51	2.37	013.13	0
013	Fat, Acid Pretreat (%)	0874	6.975	0.4300	4.463	0.6717	0.1530	51	3.74	013.02	0
013	Fat, Acid Pretreat (%)	0638	7.900	0.2000	4.463	0.6717	0.1530	51	5.12	013.10	0
015	Aluminum (mg / kg (ppm))	0297	50.00	12.00	116.9	18.51	4.741	14	-3.62	015.43	0
015	Aluminum (mg / kg (ppm))	0186	78.00	6.000	116.9	18.51	4.741	14	-2.10	015.52	0
015	Aluminum (mg / kg (ppm))	0049	102.6	5.420	116.9	18.51	4.741	14	-0.77	015.41	0
015	Aluminum (mg / kg (ppm))	0918	103.4	0.5089	116.9	18.51	4.741	14	-0.73	015.43	0
015	Aluminum (mg / kg (ppm))	0510	110.5	1.000	116.9	18.51	4.741	14	-0.35	015.43	0
015	Aluminum (mg / kg (ppm))	0407	118.9	0.7408	116.9	18.51	4.741	14	0.11	015.41	0
015	Aluminum (mg / kg (ppm))	0164	120.0	6.000	116.9	18.51	4.741	14	0.17	015.41	0
015	Aluminum (mg / kg (ppm))	0169	120.0	16.00	116.9	18.51	4.741	14	0.17	015.43	0
015	Aluminum (mg / kg (ppm))	0425	123.1	4.500	116.9	18.51	4.741	14	0.33	015.43	0
015	Aluminum (mg / kg (ppm))	0171	123.7	0.1000	116.9	18.51	4.741	14	0.36	015.41	0
015	Aluminum (mg / kg (ppm))	0560	126.0	2.100	116.9	18.51	4.741	14	0.49	015.42	0
015	Aluminum (mg / kg (ppm))	0345	130.0	2.000	116.9	18.51	4.741	14	0.71	015.43	0
015	Aluminum (mg / kg (ppm))	0520	136.0	8.000	116.9	18.51	4.741	14	1.03	015.41	0
015	Aluminum (mg / kg (ppm))	0353	156.0	2.000	116.9	18.51	4.741	14	2.11	015.43	0
017	Boron (mg / kg (ppm))	0045	5.265	0.3300	6.370	0.8903	0.4589	15	-1.24	017.42	0
017	Boron (mg / kg (ppm))	0345	5.375	0.0500	6.370	0.8903	0.4589	15	-1.12	017.43	0
017	Boron (mg / kg (ppm))	0358	5.495	1.070	6.370	0.8903	0.4589	15	-0.98	017.41	0
017	Boron (mg / kg (ppm))	0693	5.616	0.2740	6.370	0.8903	0.4589	15	-0.85	017.42	0
017	Boron (mg / kg (ppm))	0083	5.850	0.3000	6.370	0.8903	0.4589	15	-0.58	017.43	0
017	Boron (mg / kg (ppm))	0560	5.960	1.320	6.370	0.8903	0.4589	15	-0.46	017.42	0
017	Boron (mg / kg (ppm))	0407	6.193	0.1756	6.370	0.8903	0.4589	15	-0.20	017.41	0
017	Boron (mg / kg (ppm))	0918	6.410	0.0600	6.370	0.8903	0.4589	15	0.05	017.43	0
017	Boron (mg / kg (ppm))	0049	6.475	1.610	6.370	0.8903	0.4589	15	0.12	017.41	0
017	Boron (mg / kg (ppm))	0229	6.600	0.1600	6.370	0.8903	0.4589	15	0.26	017.41	0
017	Boron (mg / kg (ppm))	0226	6.760	0.0800	6.370	0.8903	0.4589	15	0.44	017.41	0
017	Boron (mg / kg (ppm))	0510	7.000	0.0000	6.370	0.8903	0.4589	15	0.71	017.43	0
017	Boron (mg / kg (ppm))	0171	7.340	0.2740	6.370	0.8903	0.4589	15	1.09	017.41	0
017	Boron (mg / kg (ppm))	0353	7.500	1.000	6.370	0.8903	0.4589	15	1.27	017.43	0
017	Boron (mg / kg (ppm))	0294	9.790	0.1800	6.370	0.8903	0.4589	15	3.84	017.42	0
017	Boron (mg / kg (ppm))	0265	10.30	2.800	6.370	0.8903	0.4589	15	4.41	017.42	1
017	Boron (mg / kg (ppm))	0297	0.0000	0.0000	6.370	0.8903	0.4589	15		017.43	4
019	Calcium (%)	2200	1.365	0.0700	2.068	0.0929	0.0444	135	-7.57	019.08	0
019	Calcium (%)	0297	1.635	0.0100	2.068	0.0929	0.0444	135	-4.66	019.43	0
019	Calcium (%)	2199	1.640	0.0800	2.068	0.0929	0.0444	135	-4.61	019.08	0
019	Calcium (%)	0720	1.770	0.0200	2.068	0.0929	0.0444	135	-3.21	019.41	0
019	Calcium (%)	0596	1.775	0.0500	2.068	0.0929	0.0444	135	-3.16	019.31	0
019	Calcium (%)	2129	1.810	0.0200	2.068	0.0929	0.0444	135	-2.78	019.42	0
019	Calcium (%)	0884	1.845	0.0300	2.068	0.0929	0.0444	135	-2.40	019.31	0
019	Calcium (%)	0171	1.895	0.0100	2.068	0.0929	0.0444	135	-1.86	019.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	0407	1.936	0.0654	2.068	0.0929	0.0444	135	-1.43	019.41	0
019	Calcium (%)	0032	1.945	0.0900	2.068	0.0929	0.0444	135	-1.33	019.43	0
019	Calcium (%)	0622	1.950	0.1600	2.068	0.0929	0.0444	135	-1.27	019.31	0
019	Calcium (%)	0037	1.952	0.0600	2.068	0.0929	0.0444	135	-1.25	019.42	0
019	Calcium (%)	2058	1.960	0.0250	2.068	0.0929	0.0444	135	-1.17	019.43	0
019	Calcium (%)	0298	1.960	0.0200	2.068	0.0929	0.0444	135	-1.17	019.41	0
019	Calcium (%)	0674	1.970	0.0000	2.068	0.0929	0.0444	135	-1.06	019.31	0
019	Calcium (%)	0144	1.970	0.0400	2.068	0.0929	0.0444	135	-1.06	019.41	0
019	Calcium (%)	0205	1.975	0.0500	2.068	0.0929	0.0444	135	-1.00	019.42	0
019	Calcium (%)	0036	1.977	0.0451	2.068	0.0929	0.0444	135	-0.98	019.43	0
019	Calcium (%)	0148	1.980	0.0600	2.068	0.0929	0.0444	135	-0.95	019.41	0
019	Calcium (%)	0870	1.981	0.1010	2.068	0.0929	0.0444	135	-0.94	019.42	0
019	Calcium (%)	0910	1.985	0.1500	2.068	0.0929	0.0444	135	-0.90	019.41	0
019	Calcium (%)	0572	1.985	0.1100	2.068	0.0929	0.0444	135	-0.90	019.53	0
019	Calcium (%)	0045	1.990	0.0400	2.068	0.0929	0.0444	135	-0.84	019.42	0
019	Calcium (%)	0536	1.990	0.0000	2.068	0.0929	0.0444	135	-0.84	019.31	0
019	Calcium (%)	0948	1.994	0.0150	2.068	0.0929	0.0444	135	-0.80	019.31	0
019	Calcium (%)	0878	1.995	0.0100	2.068	0.0929	0.0444	135	-0.79	019.41	0
019	Calcium (%)	0142	1.996	0.0508	2.068	0.0929	0.0444	135	-0.78	019.31	0
019	Calcium (%)	0038	1.999	0.0170	2.068	0.0929	0.0444	135	-0.75	019.31	0
019	Calcium (%)	0186	2.000	0.0000	2.068	0.0929	0.0444	135	-0.73	019.52	0
019	Calcium (%)	0202	2.005	0.1110	2.068	0.0929	0.0444	135	-0.69	019.43	0
019	Calcium (%)	0108	2.005	0.0500	2.068	0.0929	0.0444	135	-0.68	019.31	0
019	Calcium (%)	2126	2.005	0.0300	2.068	0.0929	0.0444	135	-0.68	019.34	0
019	Calcium (%)	0139	2.005	0.0100	2.068	0.0929	0.0444	135	-0.68	019.41	0
019	Calcium (%)	0038	2.010	0.0020	2.068	0.0929	0.0444	135	-0.63	019.43	0
019	Calcium (%)	0425	2.010	0.0400	2.068	0.0929	0.0444	135	-0.63	019.43	0
019	Calcium (%)	0510	2.010	0.0000	2.068	0.0929	0.0444	135	-0.63	019.43	0
019	Calcium (%)	2031	2.013	0.0107	2.068	0.0929	0.0444	135	-0.60	019.08	0
019	Calcium (%)	2018	2.015	0.0100	2.068	0.0929	0.0444	135	-0.57	019.31	0
019	Calcium (%)	2109	2.016	0.0990	2.068	0.0929	0.0444	135	-0.57	019.00	0
019	Calcium (%)	0154	2.018	0.0152	2.068	0.0929	0.0444	135	-0.54	019.52	0
019	Calcium (%)	2146	2.020	0.0200	2.068	0.0929	0.0444	135	-0.52	019.51	0
019	Calcium (%)	2188	2.020	0.0000	2.068	0.0929	0.0444	135	-0.52	019.08	0
019	Calcium (%)	0242	2.020	0.0400	2.068	0.0929	0.0444	135	-0.52	019.99	0
019	Calcium (%)	2006	2.021	0.0480	2.068	0.0929	0.0444	135	-0.51	019.00	0
019	Calcium (%)	0345	2.022	0.0020	2.068	0.0929	0.0444	135	-0.50	019.43	0
019	Calcium (%)	0123	2.025	0.0100	2.068	0.0929	0.0444	135	-0.47	019.41	0
019	Calcium (%)	0610	2.025	0.0300	2.068	0.0929	0.0444	135	-0.47	019.43	0
019	Calcium (%)	0878	2.030	0.0400	2.068	0.0929	0.0444	135	-0.41	019.00	0
019	Calcium (%)	2196	2.030	0.0000	2.068	0.0929	0.0444	135	-0.41	019.08	0
019	Calcium (%)	0169	2.030	0.0400	2.068	0.0929	0.0444	135	-0.41	019.32	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	0049	2.030	0.0800	2.068	0.0929	0.0444	135	-0.41	019.41	0
019	Calcium (%)	0019	2.030	0.0200	2.068	0.0929	0.0444	135	-0.41	019.41	0
019	Calcium (%)	0098	2.035	0.0700	2.068	0.0929	0.0444	135	-0.36	019.41	0
019	Calcium (%)	0265	2.040	0.0800	2.068	0.0929	0.0444	135	-0.30	019.42	0
019	Calcium (%)	0013	2.040	0.0800	2.068	0.0929	0.0444	135	-0.30	019.43	0
019	Calcium (%)	0675	2.040	0.0200	2.068	0.0929	0.0444	135	-0.30	019.43	0
019	Calcium (%)	2062	2.045	0.0239	2.068	0.0929	0.0444	135	-0.26	019.00	0
019	Calcium (%)	0676	2.045	0.1100	2.068	0.0929	0.0444	135	-0.25	019.99	0
019	Calcium (%)	0693	2.047	0.0680	2.068	0.0929	0.0444	135	-0.23	019.42	0
019	Calcium (%)	0563	2.049	0.1239	2.068	0.0929	0.0444	135	-0.21	019.31	0
019	Calcium (%)	0897	2.050	0.0200	2.068	0.0929	0.0444	135	-0.20	019.00	0
019	Calcium (%)	0849	2.050	0.0200	2.068	0.0929	0.0444	135	-0.20	019.08	0
019	Calcium (%)	0687	2.050	0.1000	2.068	0.0929	0.0444	135	-0.20	019.31	0
019	Calcium (%)	0098	2.050	0.0200	2.068	0.0929	0.0444	135	-0.20	019.43	0
019	Calcium (%)	2234	2.050	0.1400	2.068	0.0929	0.0444	135	-0.20	019.43	0
019	Calcium (%)	0357	2.052	0.0034	2.068	0.0929	0.0444	135	-0.18	019.42	0
019	Calcium (%)	0651	2.055	0.0300	2.068	0.0929	0.0444	135	-0.14	019.00	0
019	Calcium (%)	0229	2.055	0.0300	2.068	0.0929	0.0444	135	-0.14	019.41	0
019	Calcium (%)	0511	2.055	0.1100	2.068	0.0929	0.0444	135	-0.14	019.41	0
019	Calcium (%)	0848	2.055	0.0100	2.068	0.0929	0.0444	135	-0.14	019.41	0
019	Calcium (%)	0590	2.055	0.0500	2.068	0.0929	0.0444	135	-0.14	019.99	0
019	Calcium (%)	2259	2.056	0.0320	2.068	0.0929	0.0444	135	-0.13	019.41	0
019	Calcium (%)	0968	2.060	0.0070	2.068	0.0929	0.0444	135	-0.09	019.43	0
019	Calcium (%)	0893	2.060	0.0000	2.068	0.0929	0.0444	135	-0.09	019.00	0
019	Calcium (%)	2196	2.060	0.0000	2.068	0.0929	0.0444	135	-0.09	019.31	0
019	Calcium (%)	0918	2.063	0.0520	2.068	0.0929	0.0444	135	-0.06	019.43	0
019	Calcium (%)	0590	2.065	0.0100	2.068	0.0929	0.0444	135	-0.03	019.08	0
019	Calcium (%)	0186	2.065	0.1700	2.068	0.0929	0.0444	135	-0.03	019.42	0
019	Calcium (%)	2089	2.065	0.0300	2.068	0.0929	0.0444	135	-0.03	019.43	0
019	Calcium (%)	0294	2.070	0.0000	2.068	0.0929	0.0444	135	0.02	019.42	0
019	Calcium (%)	0010	2.070	0.0400	2.068	0.0929	0.0444	135	0.02	019.43	0
019	Calcium (%)	0164	2.080	0.1000	2.068	0.0929	0.0444	135	0.13	019.41	0
019	Calcium (%)	2152	2.095	0.0300	2.068	0.0929	0.0444	135	0.29	019.08	0
019	Calcium (%)	0003	2.095	0.0500	2.068	0.0929	0.0444	135	0.29	019.41	0
019	Calcium (%)	0066	2.095	0.0300	2.068	0.0929	0.0444	135	0.29	019.41	0
019	Calcium (%)	0035	2.095	0.0500	2.068	0.0929	0.0444	135	0.29	019.42	0
019	Calcium (%)	2259	2.099	0.0070	2.068	0.0929	0.0444	135	0.33	019.42	0
019	Calcium (%)	0874	2.100	0.0200	2.068	0.0929	0.0444	135	0.34	019.31	0
019	Calcium (%)	0083	2.100	0.0200	2.068	0.0929	0.0444	135	0.34	019.43	0
019	Calcium (%)	2152	2.100	0.0200	2.068	0.0929	0.0444	135	0.34	019.44	0
019	Calcium (%)	2192	2.100	0.0400	2.068	0.0929	0.0444	135	0.34	019.43	0
019	Calcium (%)	0683	2.110	0.0600	2.068	0.0929	0.0444	135	0.45	019.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	0263	2.113	0.0231	2.068	0.0929	0.0444	135	0.48	019.42	0
019	Calcium (%)	0939	2.115	0.0100	2.068	0.0929	0.0444	135	0.50	019.34	0
019	Calcium (%)	2246	2.120	0.0000	2.068	0.0929	0.0444	135	0.56	019.31	0
019	Calcium (%)	0074	2.120	0.0000	2.068	0.0929	0.0444	135	0.56	019.41	0
019	Calcium (%)	0695	2.121	0.0150	2.068	0.0929	0.0444	135	0.56	019.41	0
019	Calcium (%)	0011	2.122	0.0906	2.068	0.0929	0.0444	135	0.58	019.41	0
019	Calcium (%)	0726	2.128	0.0050	2.068	0.0929	0.0444	135	0.64	019.42	0
019	Calcium (%)	2022	2.135	0.0100	2.068	0.0929	0.0444	135	0.72	019.00	0
019	Calcium (%)	2246	2.135	0.0900	2.068	0.0929	0.0444	135	0.72	019.00	0
019	Calcium (%)	0505	2.135	0.2100	2.068	0.0929	0.0444	135	0.72	019.43	0
019	Calcium (%)	0190	2.140	0.0400	2.068	0.0929	0.0444	135	0.77	019.42	0
019	Calcium (%)	0689	2.145	0.0100	2.068	0.0929	0.0444	135	0.83	019.08	0
019	Calcium (%)	0868	2.145	0.0100	2.068	0.0929	0.0444	135	0.83	019.31	0
019	Calcium (%)	0405	2.150	0.0000	2.068	0.0929	0.0444	135	0.88	019.41	0
019	Calcium (%)	0033	2.150	0.0200	2.068	0.0929	0.0444	135	0.88	019.43	0
019	Calcium (%)	0366	2.155	0.0900	2.068	0.0929	0.0444	135	0.93	019.42	0
019	Calcium (%)	0520	2.160	0.0800	2.068	0.0929	0.0444	135	0.99	019.41	0
019	Calcium (%)	0668	2.160	0.0000	2.068	0.0929	0.0444	135	0.99	019.43	0
019	Calcium (%)	0278	2.165	0.0100	2.068	0.0929	0.0444	135	1.04	019.42	0
019	Calcium (%)	2144	2.170	0.0200	2.068	0.0929	0.0444	135	1.10	019.00	0
019	Calcium (%)	0529	2.170	0.0200	2.068	0.0929	0.0444	135	1.10	019.31	0
019	Calcium (%)	0100	2.170	0.0400	2.068	0.0929	0.0444	135	1.10	019.99	0
019	Calcium (%)	0001	2.179	0.1250	2.068	0.0929	0.0444	135	1.19	019.31	0
019	Calcium (%)	0682	2.180	0.0000	2.068	0.0929	0.0444	135	1.20	019.43	0
019	Calcium (%)	0560	2.182	0.0881	2.068	0.0929	0.0444	135	1.22	019.42	0
019	Calcium (%)	0036	2.183	0.0268	2.068	0.0929	0.0444	135	1.24	019.03	0
019	Calcium (%)	0358	2.185	0.0100	2.068	0.0929	0.0444	135	1.26	019.41	0
019	Calcium (%)	0598	2.186	0.0590	2.068	0.0929	0.0444	135	1.26	019.41	0
019	Calcium (%)	2190	2.188	0.1705	2.068	0.0929	0.0444	135	1.29	019.08	0
019	Calcium (%)	2193	2.190	0.0000	2.068	0.0929	0.0444	135	1.31	019.08	0
019	Calcium (%)	2258	2.195	0.0300	2.068	0.0929	0.0444	135	1.36	019.31	0
019	Calcium (%)	0187	2.195	0.0100	2.068	0.0929	0.0444	135	1.36	019.42	0
019	Calcium (%)	0656	2.200	0.0800	2.068	0.0929	0.0444	135	1.42	019.31	0
019	Calcium (%)	2222	2.210	0.0000	2.068	0.0929	0.0444	135	1.53	019.31	0
019	Calcium (%)	0612	2.210	0.0800	2.068	0.0929	0.0444	135	1.53	019.33	0
019	Calcium (%)	0226	2.215	0.0100	2.068	0.0929	0.0444	135	1.58	019.41	0
019	Calcium (%)	0964	2.215	0.1542	2.068	0.0929	0.0444	135	1.58	019.43	0
019	Calcium (%)	0885	2.229	0.1196	2.068	0.0929	0.0444	135	1.73	019.08	0
019	Calcium (%)	0619	2.235	0.0900	2.068	0.0929	0.0444	135	1.80	019.41	0
019	Calcium (%)	0035	2.265	0.0100	2.068	0.0929	0.0444	135	2.12	019.43	0
019	Calcium (%)	0047	2.310	0.1200	2.068	0.0929	0.0444	135	2.60	019.52	0
019	Calcium (%)	2006	2.344	0.0030	2.068	0.0929	0.0444	135	2.96	019.09	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	0353	2.385	0.0300	2.068	0.0929	0.0444	135	3.41	019.43	0
019	Calcium (%)	0852	1.650	0.3000	2.068	0.0929	0.0444	135	-4.50	019.08	1
019	Calcium (%)	0042	2.035	0.2700	2.068	0.0929	0.0444	135	-0.36	019.42	1
019	Calcium (%)	2055	2.188	0.5094	2.068	0.0929	0.0444	135	1.28	019.99	1
019	Calcium (%)	2128	1.061	0.0470	2.068	0.0929	0.0444	135	-10.85	019.02	2
021	Cobalt (mg / kg (ppm))	2234	1.100	0.0400	2.728	0.5167	0.1641	21	-3.15	021.53	0
021	Cobalt (mg / kg (ppm))	0910	1.770	0.3200	2.728	0.5167	0.1641	21	-1.85	021.52	0
021	Cobalt (mg / kg (ppm))	0619	2.030	0.0400	2.728	0.5167	0.1641	21	-1.35	021.41	0
021	Cobalt (mg / kg (ppm))	0918	2.307	0.3216	2.728	0.5167	0.1641	21	-0.82	021.53	0
021	Cobalt (mg / kg (ppm))	0047	2.350	0.1800	2.728	0.5167	0.1641	21	-0.73	021.52	0
021	Cobalt (mg / kg (ppm))	0345	2.380	0.0400	2.728	0.5167	0.1641	21	-0.67	021.43	0
021	Cobalt (mg / kg (ppm))	0693	2.416	0.2390	2.728	0.5167	0.1641	21	-0.60	021.42	0
021	Cobalt (mg / kg (ppm))	0186	2.490	0.1400	2.728	0.5167	0.1641	21	-0.46	021.52	0
021	Cobalt (mg / kg (ppm))	0560	2.745	0.0300	2.728	0.5167	0.1641	21	0.03	021.42	0
021	Cobalt (mg / kg (ppm))	0353	2.768	0.0879	2.728	0.5167	0.1641	21	0.08	021.43	0
021	Cobalt (mg / kg (ppm))	0164	2.850	0.1000	2.728	0.5167	0.1641	21	0.24	021.31	0
021	Cobalt (mg / kg (ppm))	0169	2.850	0.3400	2.728	0.5167	0.1641	21	0.24	021.43	0
021	Cobalt (mg / kg (ppm))	0171	2.855	0.0340	2.728	0.5167	0.1641	21	0.25	021.41	0
021	Cobalt (mg / kg (ppm))	0045	2.905	0.2300	2.728	0.5167	0.1641	21	0.34	021.42	0
021	Cobalt (mg / kg (ppm))	0407	2.917	0.0715	2.728	0.5167	0.1641	21	0.37	021.41	0
021	Cobalt (mg / kg (ppm))	0038	3.035	0.5450	2.728	0.5167	0.1641	21	0.59	021.43	0
021	Cobalt (mg / kg (ppm))	0148	3.140	0.0200	2.728	0.5167	0.1641	21	0.80	021.41	0
021	Cobalt (mg / kg (ppm))	0011	3.256	0.4090	2.728	0.5167	0.1641	21	1.02	021.41	0
021	Cobalt (mg / kg (ppm))	0939	3.350	0.1000	2.728	0.5167	0.1641	21	1.20	021.31	0
021	Cobalt (mg / kg (ppm))	0510	3.365	0.0100	2.728	0.5167	0.1641	21	1.23	021.43	0
021	Cobalt (mg / kg (ppm))	0964	3.375	0.1490	2.728	0.5167	0.1641	21	1.25	021.43	0
022	Copper (mg / kg (ppm))	0297	15.00	0.0000	63.56	11.03	2.338	86	-4.40	022.43	0
022	Copper (mg / kg (ppm))	2188	25.61	0.2000	63.56	11.03	2.338	86	-3.44	022.31	0
022	Copper (mg / kg (ppm))	0848	37.53	1.470	63.56	11.03	2.338	86	-2.36	022.41	0
022	Copper (mg / kg (ppm))	0598	42.46	1.610	63.56	11.03	2.338	86	-1.91	022.41	0
022	Copper (mg / kg (ppm))	0536	44.48	0.3500	63.56	11.03	2.338	86	-1.73	022.31	0
022	Copper (mg / kg (ppm))	0229	44.82	0.6000	63.56	11.03	2.338	86	-1.70	022.41	0
022	Copper (mg / kg (ppm))	0511	46.00	0.0000	63.56	11.03	2.338	86	-1.59	022.41	0
022	Copper (mg / kg (ppm))	2192	47.33	3.440	63.56	11.03	2.338	86	-1.47	022.43	0
022	Copper (mg / kg (ppm))	0619	47.45	1.500	63.56	11.03	2.338	86	-1.46	022.41	0
022	Copper (mg / kg (ppm))	0171	48.35	0.6100	63.56	11.03	2.338	86	-1.38	022.41	0
022	Copper (mg / kg (ppm))	0868	48.42	0.7900	63.56	11.03	2.338	86	-1.37	022.31	0
022	Copper (mg / kg (ppm))	0003	48.50	3.000	63.56	11.03	2.338	86	-1.37	022.41	0
022	Copper (mg / kg (ppm))	0590	48.60	2.400	63.56	11.03	2.338	86	-1.36	022.99	0
022	Copper (mg / kg (ppm))	0001	50.03	1.080	63.56	11.03	2.338	86	-1.23	022.31	0
022	Copper (mg / kg (ppm))	0948	50.87	0.9230	63.56	11.03	2.338	86	-1.15	022.33	0
022	Copper (mg / kg (ppm))	0019	50.87	10.02	63.56	11.03	2.338	86	-1.15	022.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
022	Copper (mg / kg (ppm))	0689	51.80	2.600	63.56	11.03	2.338	86	-1.07	022.31	0
022	Copper (mg / kg (ppm))	0695	51.96	1.245	63.56	11.03	2.338	86	-1.05	022.41	0
022	Copper (mg / kg (ppm))	0074	52.00	2.000	63.56	11.03	2.338	86	-1.05	022.41	0
022	Copper (mg / kg (ppm))	0049	52.53	2.700	63.56	11.03	2.338	86	-1.00	022.41	0
022	Copper (mg / kg (ppm))	0520	55.50	3.000	63.56	11.03	2.338	86	-0.73	022.41	0
022	Copper (mg / kg (ppm))	0563	55.82	3.168	63.56	11.03	2.338	86	-0.70	022.31	0
022	Copper (mg / kg (ppm))	0529	56.80	6.000	63.56	11.03	2.338	86	-0.61	022.31	0
022	Copper (mg / kg (ppm))	0720	56.90	0.3000	63.56	11.03	2.338	86	-0.60	022.41	0
022	Copper (mg / kg (ppm))	0164	57.50	7.000	63.56	11.03	2.338	86	-0.55	022.41	0
022	Copper (mg / kg (ppm))	2196	58.80	0.0000	63.56	11.03	2.338	86	-0.43	022.31	0
022	Copper (mg / kg (ppm))	2058	59.43	0.0600	63.56	11.03	2.338	86	-0.37	022.43	0
022	Copper (mg / kg (ppm))	0038	59.85	0.1000	63.56	11.03	2.338	86	-0.34	022.32	0
022	Copper (mg / kg (ppm))	0910	60.50	1.000	63.56	11.03	2.338	86	-0.28	022.41	0
022	Copper (mg / kg (ppm))	0878	62.00	4.000	63.56	11.03	2.338	86	-0.14	022.41	0
022	Copper (mg / kg (ppm))	0066	62.02	1.390	63.56	11.03	2.338	86	-0.14	022.44	0
022	Copper (mg / kg (ppm))	0656	62.44	4.530	63.56	11.03	2.338	86	-0.10	022.31	0
022	Copper (mg / kg (ppm))	0407	62.59	3.923	63.56	11.03	2.338	86	-0.09	022.41	0
022	Copper (mg / kg (ppm))	0169	62.90	4.400	63.56	11.03	2.338	86	-0.06	022.43	0
022	Copper (mg / kg (ppm))	0572	63.00	0.2000	63.56	11.03	2.338	86	-0.05	022.53	0
022	Copper (mg / kg (ppm))	0037	63.90	1.600	63.56	11.03	2.338	86	0.03	022.42	0
022	Copper (mg / kg (ppm))	0098	64.06	0.7200	63.56	11.03	2.338	86	0.05	022.41	0
022	Copper (mg / kg (ppm))	0682	64.20	0.0000	63.56	11.03	2.338	86	0.06	022.43	0
022	Copper (mg / kg (ppm))	0202	64.67	1.361	63.56	11.03	2.338	86	0.10	022.43	0
022	Copper (mg / kg (ppm))	0045	65.05	0.5000	63.56	11.03	2.338	86	0.14	022.42	0
022	Copper (mg / kg (ppm))	0968	65.50	1.000	63.56	11.03	2.338	86	0.18	022.43	0
022	Copper (mg / kg (ppm))	0726	65.76	0.0300	63.56	11.03	2.338	86	0.20	022.42	0
022	Copper (mg / kg (ppm))	0226	65.86	0.1800	63.56	11.03	2.338	86	0.21	022.41	0
022	Copper (mg / kg (ppm))	2089	66.01	5.270	63.56	11.03	2.338	86	0.22	022.43	0
022	Copper (mg / kg (ppm))	0939	66.02	0.5400	63.56	11.03	2.338	86	0.22	022.31	0
022	Copper (mg / kg (ppm))	0918	66.22	4.350	63.56	11.03	2.338	86	0.24	022.43	0
022	Copper (mg / kg (ppm))	0345	66.44	0.1100	63.56	11.03	2.338	86	0.26	022.43	0
022	Copper (mg / kg (ppm))	0870	66.49	4.580	63.56	11.03	2.338	86	0.27	022.42	0
022	Copper (mg / kg (ppm))	0505	66.92	1.080	63.56	11.03	2.338	86	0.30	022.43	0
022	Copper (mg / kg (ppm))	0042	67.45	1.100	63.56	11.03	2.338	86	0.35	022.42	0
022	Copper (mg / kg (ppm))	2246	67.85	11.60	63.56	11.03	2.338	86	0.39	022.34	0
022	Copper (mg / kg (ppm))	0035	67.86	0.0100	63.56	11.03	2.338	86	0.39	022.43	0
022	Copper (mg / kg (ppm))	0263	67.87	0.2800	63.56	11.03	2.338	86	0.39	022.42	0
022	Copper (mg / kg (ppm))	0358	67.98	2.530	63.56	11.03	2.338	86	0.40	022.41	0
022	Copper (mg / kg (ppm))	0100	68.00	0.0000	63.56	11.03	2.338	86	0.40	022.99	0
022	Copper (mg / kg (ppm))	0033	68.50	0.2000	63.56	11.03	2.338	86	0.45	022.43	0
022	Copper (mg / kg (ppm))	0186	69.00	0.0000	63.56	11.03	2.338	86	0.49	022.52	0
022	Copper (mg / kg (ppm))	0425	69.73	2.470	63.56	11.03	2.338	86	0.56	022.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
022	Copper (mg / kg (ppm))	0011	69.75	8.534	63.56	11.03	2.338	86	0.56	022.41	0
022	Copper (mg / kg (ppm))	0098	70.11	1.020	63.56	11.03	2.338	86	0.59	022.43	0
022	Copper (mg / kg (ppm))	0278	70.40	2.600	63.56	11.03	2.338	86	0.62	022.42	0
022	Copper (mg / kg (ppm))	0510	70.50	1.000	63.56	11.03	2.338	86	0.63	022.43	0
022	Copper (mg / kg (ppm))	2234	70.58	3.260	63.56	11.03	2.338	86	0.64	022.43	0
022	Copper (mg / kg (ppm))	0560	71.10	2.000	63.56	11.03	2.338	86	0.68	022.42	0
022	Copper (mg / kg (ppm))	0035	71.24	0.6900	63.56	11.03	2.338	86	0.70	022.42	0
022	Copper (mg / kg (ppm))	0013	71.35	4.700	63.56	11.03	2.338	86	0.71	022.43	0
022	Copper (mg / kg (ppm))	0032	71.50	1.400	63.56	11.03	2.338	86	0.72	022.43	0
022	Copper (mg / kg (ppm))	0610	71.50	1.000	63.56	11.03	2.338	86	0.72	022.43	0
022	Copper (mg / kg (ppm))	0668	71.75	6.100	63.56	11.03	2.338	86	0.74	022.43	0
022	Copper (mg / kg (ppm))	0693	72.85	2.591	63.56	11.03	2.338	86	0.84	022.42	0
022	Copper (mg / kg (ppm))	0205	72.90	3.000	63.56	11.03	2.338	86	0.85	022.42	0
022	Copper (mg / kg (ppm))	0366	73.00	2.000	63.56	11.03	2.338	86	0.86	022.42	0
022	Copper (mg / kg (ppm))	2129	73.08	1.520	63.56	11.03	2.338	86	0.86	022.99	0
022	Copper (mg / kg (ppm))	0242	73.50	1.000	63.56	11.03	2.338	86	0.90	022.99	0
022	Copper (mg / kg (ppm))	0190	73.76	0.0800	63.56	11.03	2.338	86	0.92	022.42	0
022	Copper (mg / kg (ppm))	0353	74.00	4.000	63.56	11.03	2.338	86	0.95	022.43	0
022	Copper (mg / kg (ppm))	0186	74.50	1.000	63.56	11.03	2.338	86	0.99	022.42	0
022	Copper (mg / kg (ppm))	0265	74.65	11.90	63.56	11.03	2.338	86	1.01	022.42	0
022	Copper (mg / kg (ppm))	0010	74.85	0.1000	63.56	11.03	2.338	86	1.02	022.33	0
022	Copper (mg / kg (ppm))	0083	75.00	4.000	63.56	11.03	2.338	86	1.04	022.43	0
022	Copper (mg / kg (ppm))	0357	75.13	0.1630	63.56	11.03	2.338	86	1.05	022.42	0
022	Copper (mg / kg (ppm))	0964	76.62	1.423	63.56	11.03	2.338	86	1.18	022.43	0
022	Copper (mg / kg (ppm))	0294	77.89	10.98	63.56	11.03	2.338	86	1.30	022.42	0
022	Copper (mg / kg (ppm))	0187	79.35	0.7000	63.56	11.03	2.338	86	1.43	022.42	0
022	Copper (mg / kg (ppm))	0675	87.33	0.6600	63.56	11.03	2.338	86	2.16	022.43	0
022	Copper (mg / kg (ppm))	0047	100.0	8.510	63.56	11.03	2.338	86	3.31	022.52	0
022	Copper (mg / kg (ppm))	0148	55.32	16.59	63.56	11.03	2.338	86	-0.75	022.41	1
022	Copper (mg / kg (ppm))	2055	74.65	17.70	63.56	11.03	2.338	86	1.01	022.99	1
022	Copper (mg / kg (ppm))	0405	302.8	0.0000	63.56	11.03	2.338	86	21.69	022.41	2
025	Iron (mg / kg (ppm))	0405	0.5550	0.0000	408.8	39.70	12.48	71	-10.28	025.41	0
025	Iron (mg / kg (ppm))	0278	74.00	6.000	408.8	39.70	12.48	71	-8.43	025.42	0
025	Iron (mg / kg (ppm))	2188	104.9	9.460	408.8	39.70	12.48	71	-7.65	025.31	0
025	Iron (mg / kg (ppm))	0265	287.5	21.00	408.8	39.70	12.48	71	-3.06	025.42	0
025	Iron (mg / kg (ppm))	0297	288.5	5.000	408.8	39.70	12.48	71	-3.03	025.43	0
025	Iron (mg / kg (ppm))	2246	298.5	0.0000	408.8	39.70	12.48	71	-2.78	025.31	0
025	Iron (mg / kg (ppm))	0598	318.7	7.800	408.8	39.70	12.48	71	-2.27	025.41	0
025	Iron (mg / kg (ppm))	2192	333.4	8.780	408.8	39.70	12.48	71	-1.90	025.43	0
025	Iron (mg / kg (ppm))	0689	348.2	11.40	408.8	39.70	12.48	71	-1.53	025.31	0
025	Iron (mg / kg (ppm))	0968	348.5	11.00	408.8	39.70	12.48	71	-1.52	025.43	0
025	Iron (mg / kg (ppm))	0226	359.5	11.83	408.8	39.70	12.48	71	-1.24	025.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
025	Iron (mg / kg (ppm))	0345	372.2	8.100	408.8	39.70	12.48	71	-0.92	025.43	0
025	Iron (mg / kg (ppm))	0619	376.5	17.00	408.8	39.70	12.48	71	-0.81	025.41	0
025	Iron (mg / kg (ppm))	0169	376.5	39.00	408.8	39.70	12.48	71	-0.81	025.43	0
025	Iron (mg / kg (ppm))	0190	377.4	8.080	408.8	39.70	12.48	71	-0.79	025.42	0
025	Iron (mg / kg (ppm))	0407	381.7	0.1615	408.8	39.70	12.48	71	-0.68	025.41	0
025	Iron (mg / kg (ppm))	0910	383.5	1.000	408.8	39.70	12.48	71	-0.64	025.41	0
025	Iron (mg / kg (ppm))	0870	384.3	17.60	408.8	39.70	12.48	71	-0.62	025.42	0
025	Iron (mg / kg (ppm))	0294	385.0	1.200	408.8	39.70	12.48	71	-0.60	025.42	0
025	Iron (mg / kg (ppm))	0720	385.2	7.000	408.8	39.70	12.48	71	-0.59	025.41	0
025	Iron (mg / kg (ppm))	0726	387.2	0.5000	408.8	39.70	12.48	71	-0.54	025.42	0
025	Iron (mg / kg (ppm))	0123	389.5	13.00	408.8	39.70	12.48	71	-0.49	025.41	0
025	Iron (mg / kg (ppm))	0874	390.5	9.000	408.8	39.70	12.48	71	-0.46	025.31	0
025	Iron (mg / kg (ppm))	0003	392.0	8.000	408.8	39.70	12.48	71	-0.42	025.41	0
025	Iron (mg / kg (ppm))	2129	394.5	11.00	408.8	39.70	12.48	71	-0.36	025.99	0
025	Iron (mg / kg (ppm))	0695	396.0	7.832	408.8	39.70	12.48	71	-0.32	025.41	0
025	Iron (mg / kg (ppm))	0263	396.6	0.0500	408.8	39.70	12.48	71	-0.31	025.42	0
025	Iron (mg / kg (ppm))	0560	399.9	42.30	408.8	39.70	12.48	71	-0.22	025.42	0
025	Iron (mg / kg (ppm))	2196	405.3	0.0000	408.8	39.70	12.48	71	-0.09	025.31	0
025	Iron (mg / kg (ppm))	0425	406.3	4.900	408.8	39.70	12.48	71	-0.06	025.43	0
025	Iron (mg / kg (ppm))	0511	407.0	24.00	408.8	39.70	12.48	71	-0.04	025.41	0
025	Iron (mg / kg (ppm))	0675	410.1	0.7300	408.8	39.70	12.48	71	0.03	025.43	0
025	Iron (mg / kg (ppm))	0505	413.2	7.300	408.8	39.70	12.48	71	0.11	025.31	0
025	Iron (mg / kg (ppm))	0878	414.0	4.000	408.8	39.70	12.48	71	0.13	025.41	0
025	Iron (mg / kg (ppm))	0520	415.5	31.00	408.8	39.70	12.48	71	0.17	025.41	0
025	Iron (mg / kg (ppm))	0042	415.5	17.00	408.8	39.70	12.48	71	0.17	025.42	0
025	Iron (mg / kg (ppm))	0171	417.5	6.700	408.8	39.70	12.48	71	0.22	025.41	0
025	Iron (mg / kg (ppm))	0038	418.5	1.000	408.8	39.70	12.48	71	0.24	025.31	0
025	Iron (mg / kg (ppm))	0353	418.5	1.000	408.8	39.70	12.48	71	0.24	025.43	0
025	Iron (mg / kg (ppm))	0083	419.0	2.000	408.8	39.70	12.48	71	0.26	025.43	0
025	Iron (mg / kg (ppm))	0693	419.2	5.830	408.8	39.70	12.48	71	0.26	025.42	0
025	Iron (mg / kg (ppm))	0510	419.5	1.000	408.8	39.70	12.48	71	0.27	025.43	0
025	Iron (mg / kg (ppm))	0164	423.0	2.000	408.8	39.70	12.48	71	0.36	025.41	0
025	Iron (mg / kg (ppm))	0563	423.2	14.61	408.8	39.70	12.48	71	0.36	025.31	0
025	Iron (mg / kg (ppm))	0037	424.0	55.10	408.8	39.70	12.48	71	0.38	025.42	0
025	Iron (mg / kg (ppm))	0045	424.5	1.000	408.8	39.70	12.48	71	0.40	025.42	0
025	Iron (mg / kg (ppm))	0098	425.9	7.300	408.8	39.70	12.48	71	0.43	025.41	0
025	Iron (mg / kg (ppm))	0868	426.8	4.490	408.8	39.70	12.48	71	0.45	025.31	0
025	Iron (mg / kg (ppm))	0242	428.5	17.00	408.8	39.70	12.48	71	0.50	025.99	0
025	Iron (mg / kg (ppm))	0049	432.0	2.700	408.8	39.70	12.48	71	0.59	025.41	0
025	Iron (mg / kg (ppm))	0013	432.0	44.00	408.8	39.70	12.48	71	0.59	025.43	0
025	Iron (mg / kg (ppm))	0682	432.6	0.0000	408.8	39.70	12.48	71	0.60	025.43	0
025	Iron (mg / kg (ppm))	0098	433.0	6.500	408.8	39.70	12.48	71	0.61	025.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
025	Iron (mg / kg (ppm))	0100	433.0	6.000	408.8	39.70	12.48	71	0.61	025.99	0
025	Iron (mg / kg (ppm))	0074	434.0	4.000	408.8	39.70	12.48	71	0.64	025.41	0
025	Iron (mg / kg (ppm))	0918	434.5	12.67	408.8	39.70	12.48	71	0.65	025.43	0
025	Iron (mg / kg (ppm))	0148	437.3	60.46	408.8	39.70	12.48	71	0.72	025.41	0
025	Iron (mg / kg (ppm))	0610	438.0	20.00	408.8	39.70	12.48	71	0.74	025.43	0
025	Iron (mg / kg (ppm))	0019	440.6	4.800	408.8	39.70	12.48	71	0.80	025.41	0
025	Iron (mg / kg (ppm))	0358	442.7	10.36	408.8	39.70	12.48	71	0.86	025.41	0
025	Iron (mg / kg (ppm))	0229	445.0	4.200	408.8	39.70	12.48	71	0.91	025.41	0
025	Iron (mg / kg (ppm))	0939	445.2	34.30	408.8	39.70	12.48	71	0.92	025.53	0
025	Iron (mg / kg (ppm))	0366	450.0	34.00	408.8	39.70	12.48	71	1.04	025.42	0
025	Iron (mg / kg (ppm))	0536	451.6	13.25	408.8	39.70	12.48	71	1.08	025.31	0
025	Iron (mg / kg (ppm))	0904	452.3	10.12	408.8	39.70	12.48	71	1.10	025.31	0
025	Iron (mg / kg (ppm))	0035	453.0	16.60	408.8	39.70	12.48	71	1.11	025.42	0
025	Iron (mg / kg (ppm))	0656	468.5	46.10	408.8	39.70	12.48	71	1.50	025.31	0
025	Iron (mg / kg (ppm))	0035	473.0	22.40	408.8	39.70	12.48	71	1.62	025.43	0
025	Iron (mg / kg (ppm))	0964	474.3	12.90	408.8	39.70	12.48	71	1.65	025.43	0
025	Iron (mg / kg (ppm))	0187	490.6	16.01	408.8	39.70	12.48	71	2.06	025.42	0
025	Iron (mg / kg (ppm))	2089	523.4	13.31	408.8	39.70	12.48	71	2.89	025.43	0
025	Iron (mg / kg (ppm))	0668	376.0	78.00	408.8	39.70	12.48	71	-0.83	025.43	1
027	Magnesium (%)	0689	0.3200	0.0000	0.3802	0.0199	0.0084	75	-3.03	027.31	0
027	Magnesium (%)	2129	0.3255	0.0250	0.3802	0.0199	0.0084	75	-2.75	027.42	0
027	Magnesium (%)	0407	0.3326	0.0053	0.3802	0.0199	0.0084	75	-2.40	027.41	0
027	Magnesium (%)	0202	0.3380	0.0400	0.3802	0.0199	0.0084	75	-2.12	027.43	0
027	Magnesium (%)	0032	0.3450	0.0300	0.3802	0.0199	0.0084	75	-1.77	027.43	0
027	Magnesium (%)	0910	0.3500	0.0200	0.3802	0.0199	0.0084	75	-1.52	027.41	0
027	Magnesium (%)	0870	0.3547	0.0049	0.3802	0.0199	0.0084	75	-1.28	027.42	0
027	Magnesium (%)	0035	0.3550	0.0300	0.3802	0.0199	0.0084	75	-1.27	027.43	0
027	Magnesium (%)	0536	0.3600	0.0000	0.3802	0.0199	0.0084	75	-1.02	027.31	0
027	Magnesium (%)	0098	0.3600	0.0000	0.3802	0.0199	0.0084	75	-1.02	027.41	0
027	Magnesium (%)	0294	0.3600	0.0000	0.3802	0.0199	0.0084	75	-1.02	027.42	0
027	Magnesium (%)	0425	0.3600	0.0000	0.3802	0.0199	0.0084	75	-1.02	027.43	0
027	Magnesium (%)	0510	0.3600	0.0000	0.3802	0.0199	0.0084	75	-1.02	027.43	0
027	Magnesium (%)	0037	0.3635	0.0070	0.3802	0.0199	0.0084	75	-0.84	027.42	0
027	Magnesium (%)	0668	0.3635	0.0050	0.3802	0.0199	0.0084	75	-0.84	027.43	0
027	Magnesium (%)	0229	0.3650	0.0100	0.3802	0.0199	0.0084	75	-0.76	027.41	0
027	Magnesium (%)	0265	0.3650	0.0100	0.3802	0.0199	0.0084	75	-0.76	027.42	0
027	Magnesium (%)	0083	0.3650	0.0100	0.3802	0.0199	0.0084	75	-0.76	027.43	0
027	Magnesium (%)	0345	0.3650	0.0020	0.3802	0.0199	0.0084	75	-0.76	027.43	0
027	Magnesium (%)	2192	0.3650	0.0100	0.3802	0.0199	0.0084	75	-0.76	027.43	0
027	Magnesium (%)	0868	0.3700	0.0000	0.3802	0.0199	0.0084	75	-0.51	027.31	0
027	Magnesium (%)	0049	0.3700	0.0000	0.3802	0.0199	0.0084	75	-0.51	027.41	0
027	Magnesium (%)	0098	0.3700	0.0200	0.3802	0.0199	0.0084	75	-0.51	027.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
027	Magnesium (%)	0505	0.3700	0.0200	0.3802	0.0199	0.0084	75	-0.51	027.43	0
027	Magnesium (%)	0038	0.3715	0.0010	0.3802	0.0199	0.0084	75	-0.44	027.43	0
027	Magnesium (%)	0520	0.3735	0.0090	0.3802	0.0199	0.0084	75	-0.34	027.41	0
027	Magnesium (%)	0726	0.3738	0.0003	0.3802	0.0199	0.0084	75	-0.32	027.42	0
027	Magnesium (%)	0154	0.3740	0.0265	0.3802	0.0199	0.0084	75	-0.31	027.52	0
027	Magnesium (%)	0171	0.3745	0.0010	0.3802	0.0199	0.0084	75	-0.29	027.41	0
027	Magnesium (%)	0878	0.3750	0.0100	0.3802	0.0199	0.0084	75	-0.26	027.41	0
027	Magnesium (%)	0610	0.3750	0.0100	0.3802	0.0199	0.0084	75	-0.26	027.43	0
027	Magnesium (%)	0590	0.3750	0.0100	0.3802	0.0199	0.0084	75	-0.26	027.99	0
027	Magnesium (%)	0968	0.3755	0.0070	0.3802	0.0199	0.0084	75	-0.24	027.43	0
027	Magnesium (%)	0572	0.3770	0.0000	0.3802	0.0199	0.0084	75	-0.16	027.53	0
027	Magnesium (%)	0186	0.3780	0.0120	0.3802	0.0199	0.0084	75	-0.11	027.52	0
027	Magnesium (%)	0045	0.3785	0.0090	0.3802	0.0199	0.0084	75	-0.09	027.42	0
027	Magnesium (%)	0142	0.3788	0.0001	0.3802	0.0199	0.0084	75	-0.07	027.31	0
027	Magnesium (%)	0019	0.3800	0.0200	0.3802	0.0199	0.0084	75	-0.01	027.41	0
027	Magnesium (%)	0035	0.3800	0.0000	0.3802	0.0199	0.0084	75	-0.01	027.42	0
027	Magnesium (%)	0242	0.3800	0.0000	0.3802	0.0199	0.0084	75	-0.01	027.99	0
027	Magnesium (%)	0693	0.3815	0.0050	0.3802	0.0199	0.0084	75	0.07	027.42	0
027	Magnesium (%)	0074	0.3820	0.0020	0.3802	0.0199	0.0084	75	0.09	027.41	0
027	Magnesium (%)	0164	0.3820	0.0040	0.3802	0.0199	0.0084	75	0.09	027.41	0
027	Magnesium (%)	0874	0.3845	0.0010	0.3802	0.0199	0.0084	75	0.22	027.31	0
027	Magnesium (%)	0003	0.3850	0.0100	0.3802	0.0199	0.0084	75	0.24	027.41	0
027	Magnesium (%)	0190	0.3850	0.0100	0.3802	0.0199	0.0084	75	0.24	027.42	0
027	Magnesium (%)	0139	0.3858	0.0055	0.3802	0.0199	0.0084	75	0.28	027.41	0
027	Magnesium (%)	0598	0.3874	0.0072	0.3802	0.0199	0.0084	75	0.36	027.41	0
027	Magnesium (%)	0563	0.3885	0.0034	0.3802	0.0199	0.0084	75	0.42	027.31	0
027	Magnesium (%)	0656	0.3900	0.0000	0.3802	0.0199	0.0084	75	0.49	027.31	0
027	Magnesium (%)	0169	0.3900	0.0200	0.3802	0.0199	0.0084	75	0.49	027.32	0
027	Magnesium (%)	0123	0.3900	0.0000	0.3802	0.0199	0.0084	75	0.49	027.41	0
027	Magnesium (%)	0358	0.3900	0.0000	0.3802	0.0199	0.0084	75	0.49	027.41	0
027	Magnesium (%)	0186	0.3900	0.0060	0.3802	0.0199	0.0084	75	0.49	027.42	0
027	Magnesium (%)	0675	0.3900	0.0000	0.3802	0.0199	0.0084	75	0.49	027.43	0
027	Magnesium (%)	2089	0.3900	0.0000	0.3802	0.0199	0.0084	75	0.49	027.43	0
027	Magnesium (%)	0405	0.3910	0.0000	0.3802	0.0199	0.0084	75	0.54	027.41	0
027	Magnesium (%)	0187	0.3944	0.0024	0.3802	0.0199	0.0084	75	0.71	027.42	0
027	Magnesium (%)	0695	0.3945	0.0030	0.3802	0.0199	0.0084	75	0.72	027.41	0
027	Magnesium (%)	0366	0.3950	0.0300	0.3802	0.0199	0.0084	75	0.74	027.42	0
027	Magnesium (%)	0918	0.3970	0.0080	0.3802	0.0199	0.0084	75	0.85	027.43	0
027	Magnesium (%)	2196	0.4000	0.0000	0.3802	0.0199	0.0084	75	1.00	027.31	0
027	Magnesium (%)	0720	0.4000	0.0200	0.3802	0.0199	0.0084	75	1.00	027.41	0
027	Magnesium (%)	0357	0.4023	0.0056	0.3802	0.0199	0.0084	75	1.11	027.42	0
027	Magnesium (%)	0560	0.4025	0.0058	0.3802	0.0199	0.0084	75	1.12	027.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
027	Magnesium (%)	0263	0.4032	0.0088	0.3802	0.0199	0.0084	75	1.16	027.42	0
027	Magnesium (%)	0226	0.4050	0.0100	0.3802	0.0199	0.0084	75	1.25	027.41	0
027	Magnesium (%)	0948	0.4070	0.0020	0.3802	0.0199	0.0084	75	1.35	027.33	0
027	Magnesium (%)	0619	0.4070	0.0060	0.3802	0.0199	0.0084	75	1.35	027.41	0
027	Magnesium (%)	0278	0.4100	0.0200	0.3802	0.0199	0.0084	75	1.50	027.42	0
027	Magnesium (%)	0964	0.4153	0.0282	0.3802	0.0199	0.0084	75	1.77	027.43	0
027	Magnesium (%)	0353	0.4200	0.0200	0.3802	0.0199	0.0084	75	2.00	027.43	0
027	Magnesium (%)	0100	0.4250	0.0100	0.3802	0.0199	0.0084	75	2.25	027.99	0
027	Magnesium (%)	0682	0.4253	0.0000	0.3802	0.0199	0.0084	75	2.27	027.43	0
027	Magnesium (%)	2246	0.4338	0.0087	0.3802	0.0199	0.0084	75	2.69	027.31	0
027	Magnesium (%)	0148	0.3250	0.0700	0.3802	0.0199	0.0084	75	-2.77	027.41	1
027	Magnesium (%)	0013	0.3945	0.0690	0.3802	0.0199	0.0084	75	0.72	027.43	1
027	Magnesium (%)	0042	0.4005	0.0730	0.3802	0.0199	0.0084	75	1.02	027.42	1
027	Magnesium (%)	0297	0.2000	0.0000	0.3802	0.0199	0.0084	75	-9.06	027.43	2
028	Manganese (mg / kg (ppm))	2188	50.79	1.510	244.9	15.99	6.487	85	-12.14	028.31	0
028	Manganese (mg / kg (ppm))	0297	54.50	1.000	244.9	15.99	6.487	85	-11.91	028.43	0
028	Manganese (mg / kg (ppm))	2234	93.29	4.070	244.9	15.99	6.487	85	-9.48	028.43	0
028	Manganese (mg / kg (ppm))	0226	185.5	14.15	244.9	15.99	6.487	85	-3.71	028.41	0
028	Manganese (mg / kg (ppm))	0689	199.8	11.50	244.9	15.99	6.487	85	-2.82	028.31	0
028	Manganese (mg / kg (ppm))	2246	202.7	5.934	244.9	15.99	6.487	85	-2.64	028.34	0
028	Manganese (mg / kg (ppm))	2192	204.3	2.410	244.9	15.99	6.487	85	-2.54	028.53	0
028	Manganese (mg / kg (ppm))	0910	210.5	3.000	244.9	15.99	6.487	85	-2.15	028.41	0
028	Manganese (mg / kg (ppm))	2058	218.5	5.270	244.9	15.99	6.487	85	-1.65	028.43	0
028	Manganese (mg / kg (ppm))	0407	219.1	2.164	244.9	15.99	6.487	85	-1.61	028.41	0
028	Manganese (mg / kg (ppm))	0047	223.2	0.7000	244.9	15.99	6.487	85	-1.36	028.52	0
028	Manganese (mg / kg (ppm))	0164	224.0	12.00	244.9	15.99	6.487	85	-1.31	028.41	0
028	Manganese (mg / kg (ppm))	0619	224.5	7.000	244.9	15.99	6.487	85	-1.27	028.41	0
028	Manganese (mg / kg (ppm))	0720	229.6	8.590	244.9	15.99	6.487	85	-0.95	028.41	0
028	Manganese (mg / kg (ppm))	0049	230.3	1.060	244.9	15.99	6.487	85	-0.91	028.41	0
028	Manganese (mg / kg (ppm))	0148	231.7	7.550	244.9	15.99	6.487	85	-0.82	028.41	0
028	Manganese (mg / kg (ppm))	0511	232.0	16.00	244.9	15.99	6.487	85	-0.81	028.41	0
028	Manganese (mg / kg (ppm))	0520	232.0	14.00	244.9	15.99	6.487	85	-0.81	028.41	0
028	Manganese (mg / kg (ppm))	0345	233.5	3.000	244.9	15.99	6.487	85	-0.71	028.43	0
028	Manganese (mg / kg (ppm))	0870	234.9	3.400	244.9	15.99	6.487	85	-0.62	028.42	0
028	Manganese (mg / kg (ppm))	0505	234.9	11.80	244.9	15.99	6.487	85	-0.62	028.43	0
028	Manganese (mg / kg (ppm))	0695	236.1	6.284	244.9	15.99	6.487	85	-0.55	028.41	0
028	Manganese (mg / kg (ppm))	0510	236.5	1.000	244.9	15.99	6.487	85	-0.52	028.43	0
028	Manganese (mg / kg (ppm))	0563	237.4	3.854	244.9	15.99	6.487	85	-0.47	028.31	0
028	Manganese (mg / kg (ppm))	0572	237.5	3.000	244.9	15.99	6.487	85	-0.46	028.53	0
028	Manganese (mg / kg (ppm))	0074	238.0	2.000	244.9	15.99	6.487	85	-0.43	028.41	0
028	Manganese (mg / kg (ppm))	0019	238.2	7.100	244.9	15.99	6.487	85	-0.42	028.41	0
028	Manganese (mg / kg (ppm))	0263	238.9	0.9600	244.9	15.99	6.487	85	-0.37	028.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
028	Manganese (mg / kg (ppm))	0868	239.0	9.090	244.9	15.99	6.487	85	-0.37	028.31	0
028	Manganese (mg / kg (ppm))	0171	239.4	0.4000	244.9	15.99	6.487	85	-0.34	028.41	0
028	Manganese (mg / kg (ppm))	0693	239.6	7.111	244.9	15.99	6.487	85	-0.33	028.42	0
028	Manganese (mg / kg (ppm))	0032	240.0	12.00	244.9	15.99	6.487	85	-0.30	028.43	0
028	Manganese (mg / kg (ppm))	0405	240.5	0.0000	244.9	15.99	6.487	85	-0.28	028.41	0
028	Manganese (mg / kg (ppm))	0003	241.0	14.00	244.9	15.99	6.487	85	-0.24	028.41	0
028	Manganese (mg / kg (ppm))	0968	241.5	1.000	244.9	15.99	6.487	85	-0.21	028.43	0
028	Manganese (mg / kg (ppm))	0598	241.7	1.600	244.9	15.99	6.487	85	-0.20	028.41	0
028	Manganese (mg / kg (ppm))	0045	243.0	8.000	244.9	15.99	6.487	85	-0.12	028.42	0
028	Manganese (mg / kg (ppm))	0590	243.0	2.000	244.9	15.99	6.487	85	-0.12	028.99	0
028	Manganese (mg / kg (ppm))	0884	243.4	4.400	244.9	15.99	6.487	85	-0.09	028.31	0
028	Manganese (mg / kg (ppm))	0001	244.0	13.90	244.9	15.99	6.487	85	-0.06	028.31	0
028	Manganese (mg / kg (ppm))	2196	244.3	0.0000	244.9	15.99	6.487	85	-0.04	028.31	0
028	Manganese (mg / kg (ppm))	0425	244.4	3.900	244.9	15.99	6.487	85	-0.03	028.43	0
028	Manganese (mg / kg (ppm))	0033	244.5	3.000	244.9	15.99	6.487	85	-0.02	028.43	0
028	Manganese (mg / kg (ppm))	0229	246.0	5.900	244.9	15.99	6.487	85	0.07	028.41	0
028	Manganese (mg / kg (ppm))	0066	246.3	7.775	244.9	15.99	6.487	85	0.09	028.44	0
028	Manganese (mg / kg (ppm))	2129	246.5	12.50	244.9	15.99	6.487	85	0.10	028.99	0
028	Manganese (mg / kg (ppm))	0038	247.5	7.000	244.9	15.99	6.487	85	0.16	028.32	0
028	Manganese (mg / kg (ppm))	0035	247.6	0.3000	244.9	15.99	6.487	85	0.17	028.42	0
028	Manganese (mg / kg (ppm))	0265	248.0	8.000	244.9	15.99	6.487	85	0.20	028.42	0
028	Manganese (mg / kg (ppm))	0098	249.4	5.800	244.9	15.99	6.487	85	0.28	028.41	0
028	Manganese (mg / kg (ppm))	0042	250.5	3.000	244.9	15.99	6.487	85	0.35	028.42	0
028	Manganese (mg / kg (ppm))	0948	250.8	3.081	244.9	15.99	6.487	85	0.37	028.33	0
028	Manganese (mg / kg (ppm))	0878	251.0	4.000	244.9	15.99	6.487	85	0.38	028.41	0
028	Manganese (mg / kg (ppm))	0278	251.5	1.000	244.9	15.99	6.487	85	0.41	028.42	0
028	Manganese (mg / kg (ppm))	0353	252.0	4.000	244.9	15.99	6.487	85	0.45	028.43	0
028	Manganese (mg / kg (ppm))	0186	253.0	12.00	244.9	15.99	6.487	85	0.51	028.52	0
028	Manganese (mg / kg (ppm))	0011	253.2	16.43	244.9	15.99	6.487	85	0.52	028.41	0
028	Manganese (mg / kg (ppm))	0668	253.5	13.00	244.9	15.99	6.487	85	0.54	028.43	0
028	Manganese (mg / kg (ppm))	0675	253.9	14.42	244.9	15.99	6.487	85	0.57	028.43	0
028	Manganese (mg / kg (ppm))	0035	254.9	2.100	244.9	15.99	6.487	85	0.62	028.43	0
028	Manganese (mg / kg (ppm))	0038	255.3	1.100	244.9	15.99	6.487	85	0.65	028.43	0
028	Manganese (mg / kg (ppm))	0242	255.5	1.000	244.9	15.99	6.487	85	0.66	028.99	0
028	Manganese (mg / kg (ppm))	0190	256.0	0.4800	244.9	15.99	6.487	85	0.70	028.42	0
028	Manganese (mg / kg (ppm))	0098	256.3	17.70	244.9	15.99	6.487	85	0.71	028.43	0
028	Manganese (mg / kg (ppm))	0013	256.5	9.000	244.9	15.99	6.487	85	0.73	028.43	0
028	Manganese (mg / kg (ppm))	0037	256.7	23.50	244.9	15.99	6.487	85	0.74	028.42	0
028	Manganese (mg / kg (ppm))	0964	257.2	12.04	244.9	15.99	6.487	85	0.77	028.43	0
028	Manganese (mg / kg (ppm))	0726	257.3	0.3000	244.9	15.99	6.487	85	0.77	028.42	0
028	Manganese (mg / kg (ppm))	0656	257.9	15.24	244.9	15.99	6.487	85	0.81	028.31	0
028	Manganese (mg / kg (ppm))	0918	258.4	1.440	244.9	15.99	6.487	85	0.85	028.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
028	Manganese (mg / kg (ppm))	0682	259.2	0.0000	244.9	15.99	6.487	85	0.89	028.43	0
028	Manganese (mg / kg (ppm))	0202	259.3	24.39	244.9	15.99	6.487	85	0.90	028.43	0
028	Manganese (mg / kg (ppm))	0536	259.7	0.8500	244.9	15.99	6.487	85	0.93	028.31	0
028	Manganese (mg / kg (ppm))	0187	259.9	1.350	244.9	15.99	6.487	85	0.94	028.42	0
028	Manganese (mg / kg (ppm))	0083	260.5	5.000	244.9	15.99	6.487	85	0.98	028.43	0
028	Manganese (mg / kg (ppm))	0358	260.6	5.610	244.9	15.99	6.487	85	0.98	028.41	0
028	Manganese (mg / kg (ppm))	2089	263.8	2.990	244.9	15.99	6.487	85	1.18	028.43	0
028	Manganese (mg / kg (ppm))	0366	264.5	13.00	244.9	15.99	6.487	85	1.23	028.42	0
028	Manganese (mg / kg (ppm))	0939	265.8	1.330	244.9	15.99	6.487	85	1.31	028.31	0
028	Manganese (mg / kg (ppm))	0186	266.5	11.00	244.9	15.99	6.487	85	1.35	028.42	0
028	Manganese (mg / kg (ppm))	0100	266.5	11.00	244.9	15.99	6.487	85	1.35	028.99	0
028	Manganese (mg / kg (ppm))	0560	268.2	10.10	244.9	15.99	6.487	85	1.46	028.42	0
028	Manganese (mg / kg (ppm))	0610	270.5	17.00	244.9	15.99	6.487	85	1.60	028.43	0
028	Manganese (mg / kg (ppm))	0357	287.5	0.8870	244.9	15.99	6.487	85	2.67	028.42	0
028	Manganese (mg / kg (ppm))	0294	315.2	4.100	244.9	15.99	6.487	85	4.39	028.42	0
028	Manganese (mg / kg (ppm))	0169	250.0	38.00	244.9	15.99	6.487	85	0.32	028.43	1
028	Manganese (mg / kg (ppm))	2055	250.3	45.30	244.9	15.99	6.487	85	0.34	028.99	1
028	Manganese (mg / kg (ppm))	2144	611.6	3.530	244.9	15.99	6.487	85	22.93	028.31	2
031	Phosphorus (%)	0108	0.7000	0.0400	0.9083	0.0356	0.0204	130	-5.86	031.01	0
031	Phosphorus (%)	0536	0.7850	0.0100	0.9083	0.0356	0.0204	130	-3.47	031.06	0
031	Phosphorus (%)	2129	0.8325	0.0321	0.9083	0.0356	0.0204	130	-2.13	031.42	0
031	Phosphorus (%)	0689	0.8450	0.0100	0.9083	0.0356	0.0204	130	-1.78	031.01	0
031	Phosphorus (%)	0676	0.8500	0.0000	0.9083	0.0356	0.0204	130	-1.64	031.99	0
031	Phosphorus (%)	0687	0.8500	0.1000	0.9083	0.0356	0.0204	130	-1.64	031.01	0
031	Phosphorus (%)	0202	0.8540	0.0040	0.9083	0.0356	0.0204	130	-1.53	031.43	0
031	Phosphorus (%)	0171	0.8550	0.0100	0.9083	0.0356	0.0204	130	-1.50	031.41	0
031	Phosphorus (%)	0294	0.8550	0.0100	0.9083	0.0356	0.0204	130	-1.50	031.42	0
031	Phosphorus (%)	0884	0.8600	0.0000	0.9083	0.0356	0.0204	130	-1.36	031.01	0
031	Phosphorus (%)	2192	0.8600	0.0200	0.9083	0.0356	0.0204	130	-1.36	031.01	0
031	Phosphorus (%)	0278	0.8600	0.0200	0.9083	0.0356	0.0204	130	-1.36	031.42	0
031	Phosphorus (%)	0010	0.8600	0.0200	0.9083	0.0356	0.0204	130	-1.36	031.43	0
031	Phosphorus (%)	0033	0.8600	0.0400	0.9083	0.0356	0.0204	130	-1.36	031.43	0
031	Phosphorus (%)	0596	0.8650	0.0300	0.9083	0.0356	0.0204	130	-1.22	031.01	0
031	Phosphorus (%)	2018	0.8650	0.0100	0.9083	0.0356	0.0204	130	-1.22	031.01	0
031	Phosphorus (%)	2199	0.8650	0.0500	0.9083	0.0356	0.0204	130	-1.22	031.01	0
031	Phosphorus (%)	2200	0.8650	0.0100	0.9083	0.0356	0.0204	130	-1.22	031.01	0
031	Phosphorus (%)	2146	0.8650	0.0300	0.9083	0.0356	0.0204	130	-1.22	031.51	0
031	Phosphorus (%)	2146	0.8700	0.0200	0.9083	0.0356	0.0204	130	-1.08	031.01	0
031	Phosphorus (%)	0019	0.8700	0.0400	0.9083	0.0356	0.0204	130	-1.08	031.41	0
031	Phosphorus (%)	0910	0.8700	0.0400	0.9083	0.0356	0.0204	130	-1.08	031.41	0
031	Phosphorus (%)	0100	0.8700	0.0000	0.9083	0.0356	0.0204	130	-1.08	031.99	0
031	Phosphorus (%)	2246	0.8715	0.0039	0.9083	0.0356	0.0204	130	-1.04	031.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	2126	0.8750	0.0500	0.9083	0.0356	0.0204	130	-0.94	031.01	0
031	Phosphorus (%)	0154	0.8795	0.0660	0.9083	0.0356	0.0204	130	-0.81	031.52	0
031	Phosphorus (%)	0066	0.8800	0.0200	0.9083	0.0356	0.0204	130	-0.80	031.44	0
031	Phosphorus (%)	0186	0.8800	0.0340	0.9083	0.0356	0.0204	130	-0.80	031.52	0
031	Phosphorus (%)	0407	0.8833	0.0015	0.9083	0.0356	0.0204	130	-0.71	031.41	0
031	Phosphorus (%)	0870	0.8838	0.0489	0.9083	0.0356	0.0204	130	-0.69	031.42	0
031	Phosphorus (%)	0520	0.8840	0.0280	0.9083	0.0356	0.0204	130	-0.68	031.41	0
031	Phosphorus (%)	0511	0.8850	0.0100	0.9083	0.0356	0.0204	130	-0.66	031.01	0
031	Phosphorus (%)	0590	0.8850	0.0100	0.9083	0.0356	0.0204	130	-0.66	031.01	0
031	Phosphorus (%)	0123	0.8850	0.0100	0.9083	0.0356	0.0204	130	-0.66	031.41	0
031	Phosphorus (%)	0142	0.8863	0.0050	0.9083	0.0356	0.0204	130	-0.62	031.01	0
031	Phosphorus (%)	0187	0.8898	0.0094	0.9083	0.0356	0.0204	130	-0.52	031.42	0
031	Phosphorus (%)	0229	0.8900	0.0200	0.9083	0.0356	0.0204	130	-0.52	031.41	0
031	Phosphorus (%)	0242	0.8900	0.0200	0.9083	0.0356	0.0204	130	-0.52	031.99	0
031	Phosphorus (%)	0651	0.8935	0.0050	0.9083	0.0356	0.0204	130	-0.42	031.01	0
031	Phosphorus (%)	0693	0.8945	0.0010	0.9083	0.0356	0.0204	130	-0.39	031.42	0
031	Phosphorus (%)	0529	0.8950	0.0300	0.9083	0.0356	0.0204	130	-0.38	031.01	0
031	Phosphorus (%)	0948	0.8950	0.0240	0.9083	0.0356	0.0204	130	-0.38	031.01	0
031	Phosphorus (%)	0726	0.8967	0.0001	0.9083	0.0356	0.0204	130	-0.33	031.42	0
031	Phosphorus (%)	2128	0.8990	0.1900	0.9083	0.0356	0.0204	130	-0.26	031.01	0
031	Phosphorus (%)	0572	0.8990	0.0740	0.9083	0.0356	0.0204	130	-0.26	031.53	0
031	Phosphorus (%)	0874	0.8995	0.0010	0.9083	0.0356	0.0204	130	-0.25	031.01	0
031	Phosphorus (%)	0893	0.9000	0.0000	0.9083	0.0356	0.0204	130	-0.23	031.00	0
031	Phosphorus (%)	0298	0.9000	0.0200	0.9083	0.0356	0.0204	130	-0.23	031.41	0
031	Phosphorus (%)	0682	0.9000	0.0000	0.9083	0.0356	0.0204	130	-0.23	031.43	0
031	Phosphorus (%)	0695	0.9005	0.0030	0.9083	0.0356	0.0204	130	-0.22	031.41	0
031	Phosphorus (%)	0563	0.9024	0.0012	0.9083	0.0356	0.0204	130	-0.17	031.01	0
031	Phosphorus (%)	0139	0.9025	0.0050	0.9083	0.0356	0.0204	130	-0.16	031.41	0
031	Phosphorus (%)	0345	0.9030	0.0100	0.9083	0.0356	0.0204	130	-0.15	031.43	0
031	Phosphorus (%)	0939	0.9050	0.0100	0.9083	0.0356	0.0204	130	-0.09	031.01	0
031	Phosphorus (%)	2193	0.9050	0.0100	0.9083	0.0356	0.0204	130	-0.09	031.01	0
031	Phosphorus (%)	2196	0.9050	0.0000	0.9083	0.0356	0.0204	130	-0.09	031.01	0
031	Phosphorus (%)	0505	0.9050	0.0100	0.9083	0.0356	0.0204	130	-0.09	031.02	0
031	Phosphorus (%)	2053	0.9050	0.0100	0.9083	0.0356	0.0204	130	-0.09	031.02	0
031	Phosphorus (%)	0049	0.9050	0.0100	0.9083	0.0356	0.0204	130	-0.09	031.41	0
031	Phosphorus (%)	2089	0.9050	0.0100	0.9083	0.0356	0.0204	130	-0.09	031.43	0
031	Phosphorus (%)	0037	0.9075	0.0130	0.9083	0.0356	0.0204	130	-0.02	031.42	0
031	Phosphorus (%)	0098	0.9093	0.0003	0.9083	0.0356	0.0204	130	0.03	031.43	0
031	Phosphorus (%)	2259	0.9095	0.0170	0.9083	0.0356	0.0204	130	0.03	031.41	0
031	Phosphorus (%)	2190	0.9096	0.0099	0.9083	0.0356	0.0204	130	0.03	031.01	0
031	Phosphorus (%)	0897	0.9100	0.0200	0.9083	0.0356	0.0204	130	0.05	031.01	0
031	Phosphorus (%)	2152	0.9100	0.0000	0.9083	0.0356	0.0204	130	0.05	031.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	0169	0.9100	0.0200	0.9083	0.0356	0.0204	130	0.05	031.03	0
031	Phosphorus (%)	0848	0.9100	0.0000	0.9083	0.0356	0.0204	130	0.05	031.41	0
031	Phosphorus (%)	0032	0.9100	0.0400	0.9083	0.0356	0.0204	130	0.05	031.43	0
031	Phosphorus (%)	0425	0.9100	0.0000	0.9083	0.0356	0.0204	130	0.05	031.43	0
031	Phosphorus (%)	0164	0.9105	0.0070	0.9083	0.0356	0.0204	130	0.06	031.41	0
031	Phosphorus (%)	0626	0.9110	0.0020	0.9083	0.0356	0.0204	130	0.07	031.01	0
031	Phosphorus (%)	0885	0.9113	0.0068	0.9083	0.0356	0.0204	130	0.08	031.01	0
031	Phosphorus (%)	0560	0.9120	0.0342	0.9083	0.0356	0.0204	130	0.10	031.42	0
031	Phosphorus (%)	0878	0.9150	0.0100	0.9083	0.0356	0.0204	130	0.19	031.01	0
031	Phosphorus (%)	2188	0.9150	0.0100	0.9083	0.0356	0.0204	130	0.19	031.01	0
031	Phosphorus (%)	0035	0.9150	0.0100	0.9083	0.0356	0.0204	130	0.19	031.42	0
031	Phosphorus (%)	0190	0.9150	0.0100	0.9083	0.0356	0.0204	130	0.19	031.42	0
031	Phosphorus (%)	2058	0.9150	0.0100	0.9083	0.0356	0.0204	130	0.19	031.43	0
031	Phosphorus (%)	2062	0.9151	0.0030	0.9083	0.0356	0.0204	130	0.19	031.00	0
031	Phosphorus (%)	0849	0.9165	0.0070	0.9083	0.0356	0.0204	130	0.23	031.01	0
031	Phosphorus (%)	2259	0.9165	0.0050	0.9083	0.0356	0.0204	130	0.23	031.42	0
031	Phosphorus (%)	0042	0.9195	0.0750	0.9083	0.0356	0.0204	130	0.31	031.42	0
031	Phosphorus (%)	0674	0.9200	0.0000	0.9083	0.0356	0.0204	130	0.33	031.01	0
031	Phosphorus (%)	0683	0.9200	0.0200	0.9083	0.0356	0.0204	130	0.33	031.01	0
031	Phosphorus (%)	0868	0.9200	0.0000	0.9083	0.0356	0.0204	130	0.33	031.01	0
031	Phosphorus (%)	0144	0.9200	0.0000	0.9083	0.0356	0.0204	130	0.33	031.41	0
031	Phosphorus (%)	2152	0.9200	0.0200	0.9083	0.0356	0.0204	130	0.33	031.44	0
031	Phosphorus (%)	0036	0.9212	0.0102	0.9083	0.0356	0.0204	130	0.36	031.43	0
031	Phosphorus (%)	0405	0.9230	0.0000	0.9083	0.0356	0.0204	130	0.41	031.41	0
031	Phosphorus (%)	0186	0.9235	0.0310	0.9083	0.0356	0.0204	130	0.43	031.42	0
031	Phosphorus (%)	0918	0.9245	0.0190	0.9083	0.0356	0.0204	130	0.45	031.43	0
031	Phosphorus (%)	0265	0.9250	0.0300	0.9083	0.0356	0.0204	130	0.47	031.42	0
031	Phosphorus (%)	2109	0.9255	0.0090	0.9083	0.0356	0.0204	130	0.48	031.01	0
031	Phosphorus (%)	0968	0.9265	0.0050	0.9083	0.0356	0.0204	130	0.51	031.43	0
031	Phosphorus (%)	0619	0.9275	0.0110	0.9083	0.0356	0.0204	130	0.54	031.41	0
031	Phosphorus (%)	0656	0.9300	0.0400	0.9083	0.0356	0.0204	130	0.61	031.01	0
031	Phosphorus (%)	2144	0.9300	0.0200	0.9083	0.0356	0.0204	130	0.61	031.01	0
031	Phosphorus (%)	0036	0.9304	0.0023	0.9083	0.0356	0.0204	130	0.62	031.03	0
031	Phosphorus (%)	0357	0.9320	0.0044	0.9083	0.0356	0.0204	130	0.66	031.42	0
031	Phosphorus (%)	0038	0.9320	0.0000	0.9083	0.0356	0.0204	130	0.66	031.43	0
031	Phosphorus (%)	0001	0.9325	0.0050	0.9083	0.0356	0.0204	130	0.68	031.03	0
031	Phosphorus (%)	0098	0.9341	0.0253	0.9083	0.0356	0.0204	130	0.72	031.41	0
031	Phosphorus (%)	2258	0.9350	0.0300	0.9083	0.0356	0.0204	130	0.75	031.01	0
031	Phosphorus (%)	0148	0.9350	0.0100	0.9083	0.0356	0.0204	130	0.75	031.41	0
031	Phosphorus (%)	0226	0.9350	0.0500	0.9083	0.0356	0.0204	130	0.75	031.41	0
031	Phosphorus (%)	0366	0.9350	0.0300	0.9083	0.0356	0.0204	130	0.75	031.42	0
031	Phosphorus (%)	0083	0.9400	0.0200	0.9083	0.0356	0.0204	130	0.89	031.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	0205	0.9425	0.0330	0.9083	0.0356	0.0204	130	0.96	031.42	0
031	Phosphorus (%)	0353	0.9450	0.0100	0.9083	0.0356	0.0204	130	1.03	031.43	0
031	Phosphorus (%)	0358	0.9450	0.0300	0.9083	0.0356	0.0204	130	1.03	031.41	0
031	Phosphorus (%)	0610	0.9450	0.0300	0.9083	0.0356	0.0204	130	1.03	031.43	0
031	Phosphorus (%)	0668	0.9480	0.0700	0.9083	0.0356	0.0204	130	1.11	031.43	0
031	Phosphorus (%)	0510	0.9500	0.0000	0.9083	0.0356	0.0204	130	1.17	031.43	0
031	Phosphorus (%)	0045	0.9540	0.0920	0.9083	0.0356	0.0204	130	1.28	031.42	0
031	Phosphorus (%)	0720	0.9550	0.0100	0.9083	0.0356	0.0204	130	1.31	031.41	0
031	Phosphorus (%)	0074	0.9590	0.0040	0.9083	0.0356	0.0204	130	1.42	031.41	0
031	Phosphorus (%)	2022	0.9600	0.0000	0.9083	0.0356	0.0204	130	1.45	031.01	0
031	Phosphorus (%)	0852	0.9600	0.0200	0.9083	0.0356	0.0204	130	1.45	031.99	0
031	Phosphorus (%)	2006	0.9605	0.0010	0.9083	0.0356	0.0204	130	1.47	031.01	0
031	Phosphorus (%)	0964	0.9658	0.0451	0.9083	0.0356	0.0204	130	1.61	031.43	0
031	Phosphorus (%)	0263	0.9821	0.0006	0.9083	0.0356	0.0204	130	2.07	031.42	0
031	Phosphorus (%)	0035	0.9850	0.0100	0.9083	0.0356	0.0204	130	2.15	031.43	0
031	Phosphorus (%)	2234	1.000	0.1000	0.9083	0.0356	0.0204	130	2.58	031.43	0
031	Phosphorus (%)	0598	1.000	0.0237	0.9083	0.0356	0.0204	130	2.58	031.41	0
031	Phosphorus (%)	2222	1.020	0.0000	0.9083	0.0356	0.0204	130	3.14	031.01	0
031	Phosphorus (%)	2031	1.021	0.0353	0.9083	0.0356	0.0204	130	3.16	031.01	0
031	Phosphorus (%)	0011	1.037	0.0866	0.9083	0.0356	0.0204	130	3.61	031.41	0
031	Phosphorus (%)	0297	1.045	0.0100	0.9083	0.0356	0.0204	130	3.84	031.43	0
031	Phosphorus (%)	0675	1.075	0.0300	0.9083	0.0356	0.0204	130	4.68	031.43	0
031	Phosphorus (%)	0003	1.415	0.9100	0.9083	0.0356	0.0204	130	14.24	031.41	2
032	Potassium (%)	0297	0.8200	0.0000	1.004	0.0630	0.0273	82	-2.92	032.43	0
032	Potassium (%)	2192	0.8250	0.0300	1.004	0.0630	0.0273	82	-2.84	032.43	0
032	Potassium (%)	0536	0.8300	0.0000	1.004	0.0630	0.0273	82	-2.76	032.31	0
032	Potassium (%)	0939	0.8450	0.0100	1.004	0.0630	0.0273	82	-2.52	032.31	0
032	Potassium (%)	0720	0.8450	0.0100	1.004	0.0630	0.0273	82	-2.52	032.41	0
032	Potassium (%)	0229	0.8950	0.0100	1.004	0.0630	0.0273	82	-1.73	032.41	0
032	Potassium (%)	0123	0.9050	0.0100	1.004	0.0630	0.0273	82	-1.57	032.41	0
032	Potassium (%)	0019	0.9100	0.0800	1.004	0.0630	0.0273	82	-1.49	032.41	0
032	Potassium (%)	0013	0.9145	0.0490	1.004	0.0630	0.0273	82	-1.42	032.43	0
032	Potassium (%)	0619	0.9195	0.0010	1.004	0.0630	0.0273	82	-1.34	032.41	0
032	Potassium (%)	0148	0.9200	0.0400	1.004	0.0630	0.0273	82	-1.33	032.41	0
032	Potassium (%)	0425	0.9250	0.0100	1.004	0.0630	0.0273	82	-1.25	032.43	0
032	Potassium (%)	0003	0.9300	0.0000	1.004	0.0630	0.0273	82	-1.17	032.41	0
032	Potassium (%)	2129	0.9332	0.0411	1.004	0.0630	0.0273	82	-1.12	032.42	0
032	Potassium (%)	0910	0.9350	0.0300	1.004	0.0630	0.0273	82	-1.09	032.41	0
032	Potassium (%)	0511	0.9400	0.1000	1.004	0.0630	0.0273	82	-1.01	032.41	0
032	Potassium (%)	0154	0.9447	0.0154	1.004	0.0630	0.0273	82	-0.94	032.52	0
032	Potassium (%)	0868	0.9500	0.0200	1.004	0.0630	0.0273	82	-0.85	032.31	0
032	Potassium (%)	2246	0.9587	0.0089	1.004	0.0630	0.0273	82	-0.72	032.32	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
032	Potassium (%)	0695	0.9590	0.0160	1.004	0.0630	0.0273	82	-0.71	032.41	0
032	Potassium (%)	0171	0.9655	0.0010	1.004	0.0630	0.0273	82	-0.61	032.41	0
032	Potassium (%)	0529	0.9750	0.0100	1.004	0.0630	0.0273	82	-0.46	032.31	0
032	Potassium (%)	0505	0.9750	0.0700	1.004	0.0630	0.0273	82	-0.46	032.43	0
032	Potassium (%)	0037	0.9830	0.0420	1.004	0.0630	0.0273	82	-0.33	032.42	0
032	Potassium (%)	0032	0.9850	0.0500	1.004	0.0630	0.0273	82	-0.30	032.43	0
032	Potassium (%)	0590	0.9850	0.0100	1.004	0.0630	0.0273	82	-0.30	032.99	0
032	Potassium (%)	0407	0.9865	0.0532	1.004	0.0630	0.0273	82	-0.27	032.41	0
032	Potassium (%)	0405	0.9880	0.0000	1.004	0.0630	0.0273	82	-0.25	032.41	0
032	Potassium (%)	0968	0.9895	0.0010	1.004	0.0630	0.0273	82	-0.23	032.43	0
032	Potassium (%)	0049	0.9900	0.0200	1.004	0.0630	0.0273	82	-0.22	032.41	0
032	Potassium (%)	0510	0.9900	0.0000	1.004	0.0630	0.0273	82	-0.22	032.43	0
032	Potassium (%)	0964	0.9940	0.0492	1.004	0.0630	0.0273	82	-0.15	032.43	0
032	Potassium (%)	0035	0.9950	0.0100	1.004	0.0630	0.0273	82	-0.14	032.43	0
032	Potassium (%)	0693	0.9985	0.0230	1.004	0.0630	0.0273	82	-0.08	032.42	0
032	Potassium (%)	0948	0.9990	0.0280	1.004	0.0630	0.0273	82	-0.08	032.31	0
032	Potassium (%)	0682	1.000	0.0000	1.004	0.0630	0.0273	82	-0.06	032.43	0
032	Potassium (%)	0038	1.004	0.0030	1.004	0.0630	0.0273	82	0.00	032.43	0
032	Potassium (%)	0598	1.005	0.0282	1.004	0.0630	0.0273	82	0.02	032.41	0
032	Potassium (%)	0726	1.007	0.0030	1.004	0.0630	0.0273	82	0.04	032.42	0
032	Potassium (%)	0042	1.007	0.1060	1.004	0.0630	0.0273	82	0.05	032.42	0
032	Potassium (%)	0202	1.007	0.0980	1.004	0.0630	0.0273	82	0.05	032.43	0
032	Potassium (%)	0142	1.008	0.0293	1.004	0.0630	0.0273	82	0.07	032.31	0
032	Potassium (%)	0870	1.008	0.0255	1.004	0.0630	0.0273	82	0.07	032.42	0
032	Potassium (%)	0683	1.010	0.0400	1.004	0.0630	0.0273	82	0.10	032.31	0
032	Potassium (%)	0100	1.010	0.0000	1.004	0.0630	0.0273	82	0.10	032.99	0
032	Potassium (%)	0186	1.012	0.0360	1.004	0.0630	0.0273	82	0.13	032.52	0
032	Potassium (%)	0139	1.013	0.0050	1.004	0.0630	0.0273	82	0.14	032.41	0
032	Potassium (%)	0045	1.014	0.0720	1.004	0.0630	0.0273	82	0.16	032.42	0
032	Potassium (%)	0074	1.015	0.0100	1.004	0.0630	0.0273	82	0.18	032.41	0
032	Potassium (%)	0164	1.015	0.0100	1.004	0.0630	0.0273	82	0.18	032.41	0
032	Potassium (%)	0265	1.020	0.0200	1.004	0.0630	0.0273	82	0.26	032.42	0
032	Potassium (%)	0035	1.025	0.0100	1.004	0.0630	0.0273	82	0.34	032.42	0
032	Potassium (%)	0186	1.025	0.0300	1.004	0.0630	0.0273	82	0.34	032.42	0
032	Potassium (%)	0278	1.025	0.0300	1.004	0.0630	0.0273	82	0.34	032.42	0
032	Potassium (%)	0083	1.030	0.0400	1.004	0.0630	0.0273	82	0.42	032.43	0
032	Potassium (%)	0878	1.035	0.0100	1.004	0.0630	0.0273	82	0.50	032.41	0
032	Potassium (%)	0205	1.035	0.0100	1.004	0.0630	0.0273	82	0.50	032.42	0
032	Potassium (%)	0874	1.045	0.0100	1.004	0.0630	0.0273	82	0.65	032.31	0
032	Potassium (%)	0144	1.045	0.0100	1.004	0.0630	0.0273	82	0.65	032.41	0
032	Potassium (%)	0187	1.045	0.0100	1.004	0.0630	0.0273	82	0.65	032.42	0
032	Potassium (%)	0098	1.045	0.0300	1.004	0.0630	0.0273	82	0.65	032.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
032	Potassium (%)	0345	1.045	0.0100	1.004	0.0630	0.0273	82	0.65	032.43	0
032	Potassium (%)	0242	1.045	0.0100	1.004	0.0630	0.0273	82	0.65	032.99	0
032	Potassium (%)	0918	1.046	0.0210	1.004	0.0630	0.0273	82	0.66	032.43	0
032	Potassium (%)	0098	1.055	0.0100	1.004	0.0630	0.0273	82	0.81	032.41	0
032	Potassium (%)	0520	1.055	0.0100	1.004	0.0630	0.0273	82	0.81	032.41	0
032	Potassium (%)	0668	1.055	0.0700	1.004	0.0630	0.0273	82	0.81	032.43	0
032	Potassium (%)	2089	1.055	0.0100	1.004	0.0630	0.0273	82	0.81	032.43	0
032	Potassium (%)	0358	1.060	0.0200	1.004	0.0630	0.0273	82	0.89	032.41	0
032	Potassium (%)	0610	1.060	0.0200	1.004	0.0630	0.0273	82	0.89	032.43	0
032	Potassium (%)	0357	1.060	0.0025	1.004	0.0630	0.0273	82	0.90	032.42	0
032	Potassium (%)	0675	1.065	0.0100	1.004	0.0630	0.0273	82	0.97	032.43	0
032	Potassium (%)	2053	1.070	0.0400	1.004	0.0630	0.0273	82	1.05	032.31	0
032	Potassium (%)	0572	1.075	0.0500	1.004	0.0630	0.0273	82	1.13	032.53	0
032	Potassium (%)	0226	1.080	0.0800	1.004	0.0630	0.0273	82	1.21	032.41	0
032	Potassium (%)	0560	1.084	0.0136	1.004	0.0630	0.0273	82	1.27	032.42	0
032	Potassium (%)	0353	1.095	0.0500	1.004	0.0630	0.0273	82	1.45	032.43	0
032	Potassium (%)	0190	1.100	0.0600	1.004	0.0630	0.0273	82	1.53	032.42	0
032	Potassium (%)	0169	1.105	0.0500	1.004	0.0630	0.0273	82	1.61	032.32	0
032	Potassium (%)	0263	1.110	0.0272	1.004	0.0630	0.0273	82	1.68	032.42	0
032	Potassium (%)	0656	1.140	0.0400	1.004	0.0630	0.0273	82	2.16	032.31	0
032	Potassium (%)	0366	1.285	0.1100	1.004	0.0630	0.0273	82	4.46	032.42	0
032	Potassium (%)	0294	0.8250	0.2900	1.004	0.0630	0.0273	82	-2.84	032.42	1
033	Salt as chloride (%)	0032	0.8900	0.0600	1.489	0.0872	0.0291	72	-6.87	033.99	0
033	Salt as chloride (%)	0726	0.9000	0.1200	1.489	0.0872	0.0291	72	-6.75	033.03	0
033	Salt as chloride (%)	2033	0.9100	0.0000	1.489	0.0872	0.0291	72	-6.64	033.99	0
033	Salt as chloride (%)	0536	1.030	0.0200	1.489	0.0872	0.0291	72	-5.26	033.99	0
033	Salt as chloride (%)	0358	1.050	0.0200	1.489	0.0872	0.0291	72	-5.03	033.99	0
033	Salt as chloride (%)	0656	1.070	0.0000	1.489	0.0872	0.0291	72	-4.80	033.03	0
033	Salt as chloride (%)	0144	1.075	0.1100	1.489	0.0872	0.0291	72	-4.75	033.03	0
033	Salt as chloride (%)	0190	1.300	0.0800	1.489	0.0872	0.0291	72	-2.17	033.03	0
033	Salt as chloride (%)	0505	1.300	0.0400	1.489	0.0872	0.0291	72	-2.17	033.03	0
033	Salt as chloride (%)	0171	1.325	0.0100	1.489	0.0872	0.0291	72	-1.88	033.00	0
033	Salt as chloride (%)	0675	1.380	0.0200	1.489	0.0872	0.0291	72	-1.25	033.00	0
033	Salt as chloride (%)	0884	1.390	0.0200	1.489	0.0872	0.0291	72	-1.13	033.00	0
033	Salt as chloride (%)	0897	1.405	0.0100	1.489	0.0872	0.0291	72	-0.96	033.00	0
033	Salt as chloride (%)	0893	1.405	0.0900	1.489	0.0872	0.0291	72	-0.96	033.00	0
033	Salt as chloride (%)	2076	1.405	0.1100	1.489	0.0872	0.0291	72	-0.96	033.00	0
033	Salt as chloride (%)	0298	1.410	0.0200	1.489	0.0872	0.0291	72	-0.90	033.00	0
033	Salt as chloride (%)	2073	1.415	0.1500	1.489	0.0872	0.0291	72	-0.85	033.99	0
033	Salt as chloride (%)	0169	1.420	0.0400	1.489	0.0872	0.0291	72	-0.79	033.00	0
033	Salt as chloride (%)	2022	1.420	0.0000	1.489	0.0872	0.0291	72	-0.79	033.00	0
033	Salt as chloride (%)	2193	1.425	0.0900	1.489	0.0872	0.0291	72	-0.73	033.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
033	Salt as chloride (%)	0425	1.425	0.0300	1.489	0.0872	0.0291	72	-0.73	033.01	0
033	Salt as chloride (%)	0003	1.450	0.1000	1.489	0.0872	0.0291	72	-0.45	033.01	0
033	Salt as chloride (%)	0619	1.455	0.0100	1.489	0.0872	0.0291	72	-0.39	033.01	0
033	Salt as chloride (%)	0885	1.463	0.0070	1.489	0.0872	0.0291	72	-0.30	033.00	0
033	Salt as chloride (%)	2259	1.472	0.0070	1.489	0.0872	0.0291	72	-0.20	033.03	0
033	Salt as chloride (%)	0353	1.475	0.0300	1.489	0.0872	0.0291	72	-0.16	033.00	0
033	Salt as chloride (%)	0265	1.480	0.0200	1.489	0.0872	0.0291	72	-0.10	033.03	0
033	Salt as chloride (%)	2062	1.490	0.0151	1.489	0.0872	0.0291	72	0.01	033.00	0
033	Salt as chloride (%)	0682	1.490	0.0000	1.489	0.0872	0.0291	72	0.01	033.99	0
033	Salt as chloride (%)	0910	1.490	0.0000	1.489	0.0872	0.0291	72	0.01	033.99	0
033	Salt as chloride (%)	0504	1.495	0.0300	1.489	0.0872	0.0291	72	0.07	033.00	0
033	Salt as chloride (%)	0853	1.500	0.0000	1.489	0.0872	0.0291	72	0.13	033.05	0
033	Salt as chloride (%)	0510	1.501	0.0000	1.489	0.0872	0.0291	72	0.14	033.01	0
033	Salt as chloride (%)	0511	1.505	0.0300	1.489	0.0872	0.0291	72	0.19	033.00	0
033	Salt as chloride (%)	0878	1.505	0.0500	1.489	0.0872	0.0291	72	0.19	033.00	0
033	Salt as chloride (%)	0868	1.505	0.0100	1.489	0.0872	0.0291	72	0.19	033.05	0
033	Salt as chloride (%)	2144	1.505	0.0100	1.489	0.0872	0.0291	72	0.19	033.99	0
033	Salt as chloride (%)	0309	1.506	0.0087	1.489	0.0872	0.0291	72	0.20	033.00	0
033	Salt as chloride (%)	0045	1.510	0.0600	1.489	0.0872	0.0291	72	0.24	033.00	0
033	Salt as chloride (%)	2196	1.510	0.0000	1.489	0.0872	0.0291	72	0.24	033.00	0
033	Salt as chloride (%)	0083	1.520	0.0000	1.489	0.0872	0.0291	72	0.36	033.01	0
033	Salt as chloride (%)	0123	1.525	0.0100	1.489	0.0872	0.0291	72	0.41	033.00	0
033	Salt as chloride (%)	0689	1.525	0.0300	1.489	0.0872	0.0291	72	0.41	033.05	0
033	Salt as chloride (%)	0098	1.530	0.0200	1.489	0.0872	0.0291	72	0.47	033.01	0
033	Salt as chloride (%)	0229	1.530	0.0000	1.489	0.0872	0.0291	72	0.47	033.01	0
033	Salt as chloride (%)	0650	1.530	0.0200	1.489	0.0872	0.0291	72	0.47	033.01	0
033	Salt as chloride (%)	2146	1.530	0.0000	1.489	0.0872	0.0291	72	0.47	033.01	0
033	Salt as chloride (%)	0164	1.535	0.0100	1.489	0.0872	0.0291	72	0.53	033.01	0
033	Salt as chloride (%)	0001	1.540	0.0000	1.489	0.0872	0.0291	72	0.59	033.01	0
033	Salt as chloride (%)	0939	1.540	0.0200	1.489	0.0872	0.0291	72	0.59	033.01	0
033	Salt as chloride (%)	0539	1.545	0.1300	1.489	0.0872	0.0291	72	0.64	033.00	0
033	Salt as chloride (%)	0100	1.545	0.0500	1.489	0.0872	0.0291	72	0.64	033.01	0
033	Salt as chloride (%)	0205	1.545	0.0100	1.489	0.0872	0.0291	72	0.64	033.01	0
033	Salt as chloride (%)	0590	1.545	0.0700	1.489	0.0872	0.0291	72	0.64	033.01	0
033	Salt as chloride (%)	0683	1.545	0.0100	1.489	0.0872	0.0291	72	0.64	033.01	0
033	Salt as chloride (%)	0964	1.550	0.0350	1.489	0.0872	0.0291	72	0.70	033.01	0
033	Salt as chloride (%)	0004	1.550	0.0200	1.489	0.0872	0.0291	72	0.70	033.01	0
033	Salt as chloride (%)	0242	1.550	0.0000	1.489	0.0872	0.0291	72	0.70	033.01	0
033	Salt as chloride (%)	0948	1.550	0.0000	1.489	0.0872	0.0291	72	0.70	033.01	0
033	Salt as chloride (%)	2043	1.550	0.0200	1.489	0.0872	0.0291	72	0.70	033.05	0
033	Salt as chloride (%)	0148	1.555	0.0100	1.489	0.0872	0.0291	72	0.76	033.01	0
033	Salt as chloride (%)	0226	1.565	0.0100	1.489	0.0872	0.0291	72	0.87	033.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
033	Salt as chloride (%)	0407	1.567	0.0010	1.489	0.0872	0.0291	72	0.89	033.01	0
033	Salt as chloride (%)	0366	1.570	0.0200	1.489	0.0872	0.0291	72	0.93	033.00	0
033	Salt as chloride (%)	2190	1.576	0.0734	1.489	0.0872	0.0291	72	1.00	033.00	0
033	Salt as chloride (%)	0019	1.580	0.0200	1.489	0.0872	0.0291	72	1.05	033.01	0
033	Salt as chloride (%)	0874	1.580	0.0000	1.489	0.0872	0.0291	72	1.05	033.01	0
033	Salt as chloride (%)	2109	1.593	0.0380	1.489	0.0872	0.0291	72	1.19	033.01	0
033	Salt as chloride (%)	0638	1.600	0.0000	1.489	0.0872	0.0291	72	1.27	033.01	0
033	Salt as chloride (%)	0644	1.611	0.0000	1.489	0.0872	0.0291	72	1.40	033.01	0
033	Salt as chloride (%)	0186	1.615	0.0300	1.489	0.0872	0.0291	72	1.45	033.01	0
033	Salt as chloride (%)	0278	1.675	0.0100	1.489	0.0872	0.0291	72	2.14	033.01	0
033	Salt as chloride (%)	0693	1.607	0.1830	1.489	0.0872	0.0291	72	1.35	033.00	1
034	Selenium (mg / kg (ppm))	0011	1.641	0.1300	2.560	0.4424	0.1484	22	-2.08	034.41	0
034	Selenium (mg / kg (ppm))	2234	1.925	0.0700	2.560	0.4424	0.1484	22	-1.43	034.53	0
034	Selenium (mg / kg (ppm))	0032	2.055	0.0300	2.560	0.4424	0.1484	22	-1.14	034.53	0
034	Selenium (mg / kg (ppm))	0910	2.145	0.4700	2.560	0.4424	0.1484	22	-0.94	034.52	0
034	Selenium (mg / kg (ppm))	0619	2.185	0.1300	2.560	0.4424	0.1484	22	-0.85	034.41	0
034	Selenium (mg / kg (ppm))	0563	2.290	0.1580	2.560	0.4424	0.1484	22	-0.61	034.04	0
034	Selenium (mg / kg (ppm))	0186	2.320	0.0600	2.560	0.4424	0.1484	22	-0.54	034.52	0
034	Selenium (mg / kg (ppm))	0171	2.420	0.3600	2.560	0.4424	0.1484	22	-0.32	034.99	0
034	Selenium (mg / kg (ppm))	0164	2.430	0.0800	2.560	0.4424	0.1484	22	-0.29	034.53	0
034	Selenium (mg / kg (ppm))	0045	2.450	0.5800	2.560	0.4424	0.1484	22	-0.25	034.04	0
034	Selenium (mg / kg (ppm))	0038	2.486	0.1290	2.560	0.4424	0.1484	22	-0.17	034.01	0
034	Selenium (mg / kg (ppm))	0047	2.530	0.3200	2.560	0.4424	0.1484	22	-0.07	034.52	0
034	Selenium (mg / kg (ppm))	0169	2.675	0.1100	2.560	0.4424	0.1484	22	0.26	034.04	0
034	Selenium (mg / kg (ppm))	0560	2.699	0.0210	2.560	0.4424	0.1484	22	0.31	034.53	0
034	Selenium (mg / kg (ppm))	0939	2.700	0.0200	2.560	0.4424	0.1484	22	0.32	034.31	0
034	Selenium (mg / kg (ppm))	0964	2.867	0.1293	2.560	0.4424	0.1484	22	0.70	034.43	0
034	Selenium (mg / kg (ppm))	2058	2.920	0.0400	2.560	0.4424	0.1484	22	0.81	034.53	0
034	Selenium (mg / kg (ppm))	0407	2.940	0.2400	2.560	0.4424	0.1484	22	0.86	034.41	0
034	Selenium (mg / kg (ppm))	2055	3.000	0.0000	2.560	0.4424	0.1484	22	1.00	034.99	0
034	Selenium (mg / kg (ppm))	0918	3.007	0.0641	2.560	0.4424	0.1484	22	1.01	034.53	0
034	Selenium (mg / kg (ppm))	0968	3.147	0.1039	2.560	0.4424	0.1484	22	1.33	034.43	0
034	Selenium (mg / kg (ppm))	0190	4.050	0.0200	2.560	0.4424	0.1484	22	3.37	034.99	0
035	Sodium (%)	0536	0.2900	0.0000	0.3698	0.0259	0.0090	85	-3.08	035.31	0
035	Sodium (%)	2192	0.2900	0.0000	0.3698	0.0259	0.0090	85	-3.08	035.43	0
035	Sodium (%)	2246	0.3201	0.0074	0.3698	0.0259	0.0090	85	-1.92	035.31	0
035	Sodium (%)	0683	0.3250	0.0100	0.3698	0.0259	0.0090	85	-1.73	035.31	0
035	Sodium (%)	0298	0.3300	0.0200	0.3698	0.0259	0.0090	85	-1.54	035.41	0
035	Sodium (%)	0171	0.3350	0.0100	0.3698	0.0259	0.0090	85	-1.34	035.41	0
035	Sodium (%)	0294	0.3400	0.0200	0.3698	0.0259	0.0090	85	-1.15	035.42	0
035	Sodium (%)	0689	0.3400	0.0000	0.3698	0.0259	0.0090	85	-1.15	035.31	0
035	Sodium (%)	0229	0.3400	0.0000	0.3698	0.0259	0.0090	85	-1.15	035.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
035	Sodium (%)	0505	0.3400	0.0000	0.3698	0.0259	0.0090	85	-1.15	035.43	0
035	Sodium (%)	0695	0.3425	0.0050	0.3698	0.0259	0.0090	85	-1.05	035.41	0
035	Sodium (%)	0619	0.3430	0.0020	0.3698	0.0259	0.0090	85	-1.03	035.41	0
035	Sodium (%)	0598	0.3438	0.0014	0.3698	0.0259	0.0090	85	-1.00	035.41	0
035	Sodium (%)	0910	0.3450	0.0100	0.3698	0.0259	0.0090	85	-0.96	035.41	0
035	Sodium (%)	0142	0.3452	0.0141	0.3698	0.0259	0.0090	85	-0.95	035.31	0
035	Sodium (%)	2006	0.3460	0.0000	0.3698	0.0259	0.0090	85	-0.92	035.01	0
035	Sodium (%)	2018	0.3500	0.0000	0.3698	0.0259	0.0090	85	-0.76	035.31	0
035	Sodium (%)	0148	0.3500	0.0000	0.3698	0.0259	0.0090	85	-0.76	035.41	0
035	Sodium (%)	0045	0.3510	0.0160	0.3698	0.0259	0.0090	85	-0.72	035.42	0
035	Sodium (%)	0868	0.3540	0.0020	0.3698	0.0259	0.0090	85	-0.61	035.01	0
035	Sodium (%)	2126	0.3550	0.0100	0.3698	0.0259	0.0090	85	-0.57	035.31	0
035	Sodium (%)	0019	0.3550	0.0100	0.3698	0.0259	0.0090	85	-0.57	035.41	0
035	Sodium (%)	0035	0.3550	0.0100	0.3698	0.0259	0.0090	85	-0.57	035.43	0
035	Sodium (%)	0297	0.3550	0.0100	0.3698	0.0259	0.0090	85	-0.57	035.43	0
035	Sodium (%)	2129	0.3556	0.0205	0.3698	0.0259	0.0090	85	-0.55	035.99	0
035	Sodium (%)	0366	0.3565	0.0070	0.3698	0.0259	0.0090	85	-0.51	035.42	0
035	Sodium (%)	0874	0.3570	0.0080	0.3698	0.0259	0.0090	85	-0.49	035.31	0
035	Sodium (%)	0563	0.3576	0.0211	0.3698	0.0259	0.0090	85	-0.47	035.31	0
035	Sodium (%)	0948	0.3580	0.0080	0.3698	0.0259	0.0090	85	-0.45	035.31	0
035	Sodium (%)	0205	0.3580	0.0140	0.3698	0.0259	0.0090	85	-0.45	035.42	0
035	Sodium (%)	0693	0.3580	0.0080	0.3698	0.0259	0.0090	85	-0.45	035.42	0
035	Sodium (%)	0345	0.3590	0.0020	0.3698	0.0259	0.0090	85	-0.42	035.43	0
035	Sodium (%)	0572	0.3595	0.0010	0.3698	0.0259	0.0090	85	-0.40	035.53	0
035	Sodium (%)	2196	0.3600	0.0000	0.3698	0.0259	0.0090	85	-0.38	035.31	0
035	Sodium (%)	2222	0.3600	0.0000	0.3698	0.0259	0.0090	85	-0.38	035.31	0
035	Sodium (%)	0049	0.3600	0.0200	0.3698	0.0259	0.0090	85	-0.38	035.41	0
035	Sodium (%)	0123	0.3600	0.0000	0.3698	0.0259	0.0090	85	-0.38	035.41	0
035	Sodium (%)	0265	0.3600	0.0000	0.3698	0.0259	0.0090	85	-0.38	035.42	0
035	Sodium (%)	0870	0.3605	0.0009	0.3698	0.0259	0.0090	85	-0.36	035.42	0
035	Sodium (%)	0035	0.3650	0.0100	0.3698	0.0259	0.0090	85	-0.18	035.42	0
035	Sodium (%)	0425	0.3650	0.0100	0.3698	0.0259	0.0090	85	-0.18	035.43	0
035	Sodium (%)	0164	0.3660	0.0000	0.3698	0.0259	0.0090	85	-0.15	035.41	0
035	Sodium (%)	0520	0.3664	0.0143	0.3698	0.0259	0.0090	85	-0.13	035.41	0
035	Sodium (%)	0202	0.3665	0.0190	0.3698	0.0259	0.0090	85	-0.13	035.43	0
035	Sodium (%)	0407	0.3682	0.0051	0.3698	0.0259	0.0090	85	-0.06	035.41	0
035	Sodium (%)	0868	0.3700	0.0000	0.3698	0.0259	0.0090	85	0.01	035.31	0
035	Sodium (%)	0510	0.3700	0.0000	0.3698	0.0259	0.0090	85	0.01	035.43	0
035	Sodium (%)	2089	0.3700	0.0200	0.3698	0.0259	0.0090	85	0.01	035.43	0
035	Sodium (%)	0590	0.3700	0.0000	0.3698	0.0259	0.0090	85	0.01	035.99	0
035	Sodium (%)	0144	0.3750	0.0100	0.3698	0.0259	0.0090	85	0.20	035.41	0
035	Sodium (%)	0190	0.3750	0.0100	0.3698	0.0259	0.0090	85	0.20	035.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
035	Sodium (%)	0726	0.3769	0.0010	0.3698	0.0259	0.0090	85	0.28	035.42	0
035	Sodium (%)	0154	0.3786	0.0210	0.3698	0.0259	0.0090	85	0.34	035.52	0
035	Sodium (%)	0884	0.3800	0.0000	0.3698	0.0259	0.0090	85	0.40	035.02	0
035	Sodium (%)	0939	0.3800	0.0200	0.3698	0.0259	0.0090	85	0.40	035.31	0
035	Sodium (%)	0098	0.3800	0.0000	0.3698	0.0259	0.0090	85	0.40	035.43	0
035	Sodium (%)	0682	0.3800	0.0000	0.3698	0.0259	0.0090	85	0.40	035.43	0
035	Sodium (%)	0242	0.3800	0.0000	0.3698	0.0259	0.0090	85	0.40	035.99	0
035	Sodium (%)	0668	0.3820	0.0400	0.3698	0.0259	0.0090	85	0.47	035.43	0
035	Sodium (%)	0098	0.3850	0.0100	0.3698	0.0259	0.0090	85	0.59	035.41	0
035	Sodium (%)	0720	0.3850	0.0100	0.3698	0.0259	0.0090	85	0.59	035.41	0
035	Sodium (%)	0878	0.3850	0.0100	0.3698	0.0259	0.0090	85	0.59	035.41	0
035	Sodium (%)	0083	0.3850	0.0100	0.3698	0.0259	0.0090	85	0.59	035.43	0
035	Sodium (%)	0037	0.3885	0.0250	0.3698	0.0259	0.0090	85	0.72	035.05	0
035	Sodium (%)	0918	0.3900	0.0160	0.3698	0.0259	0.0090	85	0.78	035.43	0
035	Sodium (%)	0100	0.3900	0.0000	0.3698	0.0259	0.0090	85	0.78	035.99	0
035	Sodium (%)	0042	0.3920	0.0300	0.3698	0.0259	0.0090	85	0.86	035.42	0
035	Sodium (%)	0186	0.3930	0.0140	0.3698	0.0259	0.0090	85	0.90	035.52	0
035	Sodium (%)	0169	0.3950	0.0100	0.3698	0.0259	0.0090	85	0.97	035.32	0
035	Sodium (%)	0610	0.3950	0.0100	0.3698	0.0259	0.0090	85	0.97	035.43	0
035	Sodium (%)	0675	0.3950	0.0100	0.3698	0.0259	0.0090	85	0.97	035.43	0
035	Sodium (%)	0187	0.3957	0.0018	0.3698	0.0259	0.0090	85	1.00	035.42	0
035	Sodium (%)	0226	0.4000	0.0200	0.3698	0.0259	0.0090	85	1.17	035.41	0
035	Sodium (%)	0186	0.4000	0.0240	0.3698	0.0259	0.0090	85	1.17	035.42	0
035	Sodium (%)	0353	0.4001	0.0124	0.3698	0.0259	0.0090	85	1.17	035.43	0
035	Sodium (%)	0263	0.4019	0.0142	0.3698	0.0259	0.0090	85	1.24	035.42	0
035	Sodium (%)	0560	0.4046	0.0078	0.3698	0.0259	0.0090	85	1.34	035.42	0
035	Sodium (%)	0278	0.4050	0.0100	0.3698	0.0259	0.0090	85	1.36	035.42	0
035	Sodium (%)	0139	0.4125	0.0050	0.3698	0.0259	0.0090	85	1.65	035.41	0
035	Sodium (%)	0656	0.4150	0.0100	0.3698	0.0259	0.0090	85	1.75	035.31	0
035	Sodium (%)	0358	0.4150	0.0100	0.3698	0.0259	0.0090	85	1.75	035.41	0
035	Sodium (%)	0964	0.4211	0.0010	0.3698	0.0259	0.0090	85	1.98	035.43	0
035	Sodium (%)	2258	0.4250	0.0100	0.3698	0.0259	0.0090	85	2.13	035.31	0
035	Sodium (%)	0038	0.4450	0.0080	0.3698	0.0259	0.0090	85	2.90	035.43	0
035	Sodium (%)	0108	0.4650	0.0300	0.3698	0.0259	0.0090	85	3.68	035.05	0
035	Sodium (%)	2055	0.4207	0.0705	0.3698	0.0259	0.0090	85	1.96	035.99	1
035	Sodium (%)	0405	0.0007	0.0000	0.3698	0.0259	0.0090	85	-14.25	035.41	2
036	Sulfur (%)	0297	0.2400	0.0200	0.3339	0.0198	0.0100	42	-4.75	036.43	0
036	Sulfur (%)	0045	0.2995	0.0050	0.3339	0.0198	0.0100	42	-1.74	036.42	0
036	Sulfur (%)	0870	0.2997	0.0046	0.3339	0.0198	0.0100	42	-1.73	036.42	0
036	Sulfur (%)	0186	0.3055	0.0050	0.3339	0.0198	0.0100	42	-1.44	036.52	0
036	Sulfur (%)	0505	0.3100	0.0400	0.3339	0.0198	0.0100	42	-1.21	036.43	0
036	Sulfur (%)	2129	0.3147	0.0096	0.3339	0.0198	0.0100	42	-0.97	036.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
036	Sulfur (%)	0229	0.3150	0.0100	0.3339	0.0198	0.0100	42	-0.96	036.42	0
036	Sulfur (%)	0202	0.3160	0.0040	0.3339	0.0198	0.0100	42	-0.91	036.43	0
036	Sulfur (%)	0366	0.3200	0.0200	0.3339	0.0198	0.0100	42	-0.70	036.42	0
036	Sulfur (%)	0169	0.3200	0.0200	0.3339	0.0198	0.0100	42	-0.70	036.43	0
036	Sulfur (%)	0345	0.3205	0.0010	0.3339	0.0198	0.0100	42	-0.68	036.43	0
036	Sulfur (%)	0726	0.3248	0.0008	0.3339	0.0198	0.0100	42	-0.46	036.42	0
036	Sulfur (%)	0229	0.3250	0.0100	0.3339	0.0198	0.0100	42	-0.45	036.04	0
036	Sulfur (%)	0164	0.3250	0.0100	0.3339	0.0198	0.0100	42	-0.45	036.42	0
036	Sulfur (%)	0035	0.3250	0.0100	0.3339	0.0198	0.0100	42	-0.45	036.43	0
036	Sulfur (%)	0693	0.3255	0.0050	0.3339	0.0198	0.0100	42	-0.43	036.42	0
036	Sulfur (%)	0265	0.3300	0.0000	0.3339	0.0198	0.0100	42	-0.20	036.42	0
036	Sulfur (%)	0242	0.3300	0.0000	0.3339	0.0198	0.0100	42	-0.20	036.99	0
036	Sulfur (%)	0186	0.3320	0.0020	0.3339	0.0198	0.0100	42	-0.10	036.42	0
036	Sulfur (%)	0910	0.3330	0.0119	0.3339	0.0198	0.0100	42	-0.05	036.52	0
036	Sulfur (%)	0407	0.3335	0.0093	0.3339	0.0198	0.0100	42	-0.02	036.43	0
036	Sulfur (%)	0357	0.3338	0.0004	0.3339	0.0198	0.0100	42	-0.01	036.42	0
036	Sulfur (%)	0187	0.3340	0.0017	0.3339	0.0198	0.0100	42	0.00	036.42	0
036	Sulfur (%)	0226	0.3350	0.0100	0.3339	0.0198	0.0100	42	0.06	036.04	0
036	Sulfur (%)	0171	0.3350	0.0100	0.3339	0.0198	0.0100	42	0.06	036.42	0
036	Sulfur (%)	0083	0.3350	0.0100	0.3339	0.0198	0.0100	42	0.06	036.43	0
036	Sulfur (%)	0098	0.3381	0.0099	0.3339	0.0198	0.0100	42	0.21	036.43	0
036	Sulfur (%)	0148	0.3400	0.0000	0.3339	0.0198	0.0100	42	0.31	036.04	0
036	Sulfur (%)	0033	0.3400	0.0000	0.3339	0.0198	0.0100	42	0.31	036.43	0
036	Sulfur (%)	0708	0.3405	0.0010	0.3339	0.0198	0.0100	42	0.33	036.42	0
036	Sulfur (%)	0668	0.3475	0.0270	0.3339	0.0198	0.0100	42	0.69	036.43	0
036	Sulfur (%)	2053	0.3500	0.0200	0.3339	0.0198	0.0100	42	0.81	036.04	0
036	Sulfur (%)	0037	0.3505	0.0150	0.3339	0.0198	0.0100	42	0.84	036.42	0
036	Sulfur (%)	0035	0.3550	0.0300	0.3339	0.0198	0.0100	42	1.07	036.42	0
036	Sulfur (%)	0353	0.3550	0.0100	0.3339	0.0198	0.0100	42	1.07	036.43	0
036	Sulfur (%)	2089	0.3550	0.0100	0.3339	0.0198	0.0100	42	1.07	036.43	0
036	Sulfur (%)	0560	0.3572	0.0154	0.3339	0.0198	0.0100	42	1.18	036.42	0
036	Sulfur (%)	0918	0.3575	0.0130	0.3339	0.0198	0.0100	42	1.19	036.43	0
036	Sulfur (%)	0190	0.3600	0.0200	0.3339	0.0198	0.0100	42	1.32	036.42	0
036	Sulfur (%)	0278	0.3600	0.0200	0.3339	0.0198	0.0100	42	1.32	036.42	0
036	Sulfur (%)	0510	0.3800	0.0000	0.3339	0.0198	0.0100	42	2.33	036.43	0
036	Sulfur (%)	0294	0.3900	0.0000	0.3339	0.0198	0.0100	42	2.84	036.42	0
036	Sulfur (%)	0032	0.3350	0.0500	0.3339	0.0198	0.0100	42	0.06	036.43	1
037	Zinc (mg / kg (ppm))	2188	39.36	1.910	285.9	29.86	11.15	87	-8.26	037.31	0
037	Zinc (mg / kg (ppm))	0297	165.5	1.000	285.9	29.86	11.15	87	-4.03	037.43	0
037	Zinc (mg / kg (ppm))	2234	196.8	9.480	285.9	29.86	11.15	87	-2.99	037.43	0
037	Zinc (mg / kg (ppm))	0047	209.5	1.900	285.9	29.86	11.15	87	-2.56	037.52	0
037	Zinc (mg / kg (ppm))	2246	212.3	3.058	285.9	29.86	11.15	87	-2.47	037.33	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
037	Zinc (mg / kg (ppm))	0848	223.2	1.160	285.9	29.86	11.15	87	-2.10	037.41	0
037	Zinc (mg / kg (ppm))	0511	237.5	15.00	285.9	29.86	11.15	87	-1.62	037.41	0
037	Zinc (mg / kg (ppm))	0695	240.4	1.759	285.9	29.86	11.15	87	-1.53	037.41	0
037	Zinc (mg / kg (ppm))	0529	242.9	8.100	285.9	29.86	11.15	87	-1.44	037.31	0
037	Zinc (mg / kg (ppm))	0910	243.0	8.000	285.9	29.86	11.15	87	-1.44	037.41	0
037	Zinc (mg / kg (ppm))	0689	243.5	21.90	285.9	29.86	11.15	87	-1.42	037.31	0
037	Zinc (mg / kg (ppm))	0619	243.5	11.00	285.9	29.86	11.15	87	-1.42	037.41	0
037	Zinc (mg / kg (ppm))	0536	249.5	0.2500	285.9	29.86	11.15	87	-1.22	037.31	0
037	Zinc (mg / kg (ppm))	0164	252.5	19.00	285.9	29.86	11.15	87	-1.12	037.41	0
037	Zinc (mg / kg (ppm))	0049	256.6	4.370	285.9	29.86	11.15	87	-0.98	037.41	0
037	Zinc (mg / kg (ppm))	2192	261.1	36.00	285.9	29.86	11.15	87	-0.83	037.43	0
037	Zinc (mg / kg (ppm))	0019	261.4	1.500	285.9	29.86	11.15	87	-0.82	037.41	0
037	Zinc (mg / kg (ppm))	0520	264.0	20.00	285.9	29.86	11.15	87	-0.73	037.41	0
037	Zinc (mg / kg (ppm))	2129	264.0	2.000	285.9	29.86	11.15	87	-0.73	037.99	0
037	Zinc (mg / kg (ppm))	2196	267.9	0.0000	285.9	29.86	11.15	87	-0.60	037.31	0
037	Zinc (mg / kg (ppm))	0878	268.5	7.000	285.9	29.86	11.15	87	-0.58	037.41	0
037	Zinc (mg / kg (ppm))	0171	268.9	4.500	285.9	29.86	11.15	87	-0.57	037.41	0
037	Zinc (mg / kg (ppm))	0598	270.7	18.40	285.9	29.86	11.15	87	-0.51	037.41	0
037	Zinc (mg / kg (ppm))	0868	271.8	1.090	285.9	29.86	11.15	87	-0.47	037.31	0
037	Zinc (mg / kg (ppm))	0720	272.6	0.9000	285.9	29.86	11.15	87	-0.45	037.41	0
037	Zinc (mg / kg (ppm))	0001	272.9	10.20	285.9	29.86	11.15	87	-0.44	037.31	0
037	Zinc (mg / kg (ppm))	0074	274.0	10.00	285.9	29.86	11.15	87	-0.40	037.41	0
037	Zinc (mg / kg (ppm))	2058	274.4	15.41	285.9	29.86	11.15	87	-0.39	037.43	0
037	Zinc (mg / kg (ppm))	0186	275.5	13.00	285.9	29.86	11.15	87	-0.35	037.42	0
037	Zinc (mg / kg (ppm))	0345	278.0	4.000	285.9	29.86	11.15	87	-0.27	037.43	0
037	Zinc (mg / kg (ppm))	0003	278.5	17.00	285.9	29.86	11.15	87	-0.25	037.41	0
037	Zinc (mg / kg (ppm))	0590	279.0	22.00	285.9	29.86	11.15	87	-0.23	037.99	0
037	Zinc (mg / kg (ppm))	0278	280.1	9.900	285.9	29.86	11.15	87	-0.20	037.42	0
037	Zinc (mg / kg (ppm))	0148	282.1	24.61	285.9	29.86	11.15	87	-0.13	037.41	0
037	Zinc (mg / kg (ppm))	0202	282.1	0.3370	285.9	29.86	11.15	87	-0.13	037.43	0
037	Zinc (mg / kg (ppm))	0038	282.5	21.00	285.9	29.86	11.15	87	-0.12	037.32	0
037	Zinc (mg / kg (ppm))	0563	282.7	17.47	285.9	29.86	11.15	87	-0.11	037.31	0
037	Zinc (mg / kg (ppm))	0948	282.8	3.145	285.9	29.86	11.15	87	-0.10	037.33	0
037	Zinc (mg / kg (ppm))	0011	283.1	2.534	285.9	29.86	11.15	87	-0.09	037.41	0
037	Zinc (mg / kg (ppm))	0939	283.7	0.2200	285.9	29.86	11.15	87	-0.08	037.31	0
037	Zinc (mg / kg (ppm))	0610	284.0	14.00	285.9	29.86	11.15	87	-0.06	037.43	0
037	Zinc (mg / kg (ppm))	0229	284.1	0.9000	285.9	29.86	11.15	87	-0.06	037.41	0
037	Zinc (mg / kg (ppm))	0098	284.8	20.10	285.9	29.86	11.15	87	-0.04	037.41	0
037	Zinc (mg / kg (ppm))	0263	285.1	2.360	285.9	29.86	11.15	87	-0.03	037.42	0
037	Zinc (mg / kg (ppm))	0186	287.0	12.00	285.9	29.86	11.15	87	0.04	037.52	0
037	Zinc (mg / kg (ppm))	0874	287.5	11.00	285.9	29.86	11.15	87	0.05	037.31	0
037	Zinc (mg / kg (ppm))	0294	287.6	14.80	285.9	29.86	11.15	87	0.06	037.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
037	Zinc (mg / kg (ppm))	0045	288.0	38.00	285.9	29.86	11.15	87	0.07	037.42	0
037	Zinc (mg / kg (ppm))	0358	289.2	16.02	285.9	29.86	11.15	87	0.11	037.41	0
037	Zinc (mg / kg (ppm))	0425	289.3	2.080	285.9	29.86	11.15	87	0.11	037.43	0
037	Zinc (mg / kg (ppm))	0042	289.5	29.00	285.9	29.86	11.15	87	0.12	037.42	0
037	Zinc (mg / kg (ppm))	0366	290.0	36.00	285.9	29.86	11.15	87	0.14	037.42	0
037	Zinc (mg / kg (ppm))	0226	290.5	3.210	285.9	29.86	11.15	87	0.15	037.41	0
037	Zinc (mg / kg (ppm))	2089	293.4	4.180	285.9	29.86	11.15	87	0.25	037.43	0
037	Zinc (mg / kg (ppm))	0407	294.1	1.452	285.9	29.86	11.15	87	0.27	037.41	0
037	Zinc (mg / kg (ppm))	0510	297.5	1.000	285.9	29.86	11.15	87	0.39	037.43	0
037	Zinc (mg / kg (ppm))	0693	299.0	17.34	285.9	29.86	11.15	87	0.44	037.42	0
037	Zinc (mg / kg (ppm))	0968	299.5	1.000	285.9	29.86	11.15	87	0.45	037.43	0
037	Zinc (mg / kg (ppm))	0572	299.5	27.00	285.9	29.86	11.15	87	0.45	037.53	0
037	Zinc (mg / kg (ppm))	0010	300.0	20.00	285.9	29.86	11.15	87	0.47	037.33	0
037	Zinc (mg / kg (ppm))	0918	301.6	0.8100	285.9	29.86	11.15	87	0.53	037.43	0
037	Zinc (mg / kg (ppm))	0187	302.9	1.300	285.9	29.86	11.15	87	0.57	037.42	0
037	Zinc (mg / kg (ppm))	0190	303.3	6.700	285.9	29.86	11.15	87	0.58	037.42	0
037	Zinc (mg / kg (ppm))	0169	303.5	41.00	285.9	29.86	11.15	87	0.59	037.43	0
037	Zinc (mg / kg (ppm))	0682	304.5	0.0000	285.9	29.86	11.15	87	0.62	037.43	0
037	Zinc (mg / kg (ppm))	0726	305.0	2.700	285.9	29.86	11.15	87	0.64	037.42	0
037	Zinc (mg / kg (ppm))	0013	306.0	56.00	285.9	29.86	11.15	87	0.67	037.43	0
037	Zinc (mg / kg (ppm))	0668	306.5	5.000	285.9	29.86	11.15	87	0.69	037.43	0
037	Zinc (mg / kg (ppm))	0505	308.7	0.5000	285.9	29.86	11.15	87	0.76	037.43	0
037	Zinc (mg / kg (ppm))	0265	309.0	10.00	285.9	29.86	11.15	87	0.77	037.42	0
037	Zinc (mg / kg (ppm))	0098	309.3	35.10	285.9	29.86	11.15	87	0.78	037.43	0
037	Zinc (mg / kg (ppm))	0353	312.0	4.000	285.9	29.86	11.15	87	0.87	037.43	0
037	Zinc (mg / kg (ppm))	0032	314.0	32.00	285.9	29.86	11.15	87	0.94	037.43	0
037	Zinc (mg / kg (ppm))	0560	317.0	7.200	285.9	29.86	11.15	87	1.04	037.42	0
037	Zinc (mg / kg (ppm))	0035	317.2	11.70	285.9	29.86	11.15	87	1.05	037.43	0
037	Zinc (mg / kg (ppm))	0037	318.9	11.10	285.9	29.86	11.15	87	1.10	037.42	0
037	Zinc (mg / kg (ppm))	0870	319.6	5.700	285.9	29.86	11.15	87	1.13	037.42	0
037	Zinc (mg / kg (ppm))	0100	322.0	2.000	285.9	29.86	11.15	87	1.21	037.99	0
037	Zinc (mg / kg (ppm))	0066	322.5	8.900	285.9	29.86	11.15	87	1.22	037.44	0
037	Zinc (mg / kg (ppm))	0083	323.5	7.000	285.9	29.86	11.15	87	1.26	037.43	0
037	Zinc (mg / kg (ppm))	0405	326.3	0.0000	285.9	29.86	11.15	87	1.35	037.41	0
037	Zinc (mg / kg (ppm))	0035	327.9	1.500	285.9	29.86	11.15	87	1.40	037.42	0
037	Zinc (mg / kg (ppm))	0357	333.2	0.4970	285.9	29.86	11.15	87	1.58	037.42	0
037	Zinc (mg / kg (ppm))	0656	334.8	15.68	285.9	29.86	11.15	87	1.64	037.31	0
037	Zinc (mg / kg (ppm))	0242	338.5	1.000	285.9	29.86	11.15	87	1.76	037.99	0
037	Zinc (mg / kg (ppm))	0964	350.7	17.76	285.9	29.86	11.15	87	2.17	037.43	0
037	Zinc (mg / kg (ppm))	0675	407.1	43.37	285.9	29.86	11.15	87	4.06	037.43	0
037	Zinc (mg / kg (ppm))	2055	299.2	84.00	285.9	29.86	11.15	87	0.44	037.99	1
037	Zinc (mg / kg (ppm))	2178	483.5	69.00	285.9	29.86	11.15	87	6.62	037.33	1

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
038	Molybdenum (mg / kg (ppm))	0910	0.8450	0.2500	1.462	0.4227	0.1394	18	-1.46	038.52	0
038	Molybdenum (mg / kg (ppm))	0353	1.028	0.2524	1.462	0.4227	0.1394	18	-1.03	038.43	0
038	Molybdenum (mg / kg (ppm))	0186	1.075	0.0900	1.462	0.4227	0.1394	18	-0.92	038.52	0
038	Molybdenum (mg / kg (ppm))	0038	1.094	0.0900	1.462	0.4227	0.1394	18	-0.87	038.43	0
038	Molybdenum (mg / kg (ppm))	0918	1.167	0.0249	1.462	0.4227	0.1394	18	-0.70	038.53	0
038	Molybdenum (mg / kg (ppm))	0345	1.253	0.0260	1.462	0.4227	0.1394	18	-0.49	038.43	0
038	Molybdenum (mg / kg (ppm))	0169	1.290	0.0200	1.462	0.4227	0.1394	18	-0.41	038.43	0
038	Molybdenum (mg / kg (ppm))	0171	1.318	0.0660	1.462	0.4227	0.1394	18	-0.34	038.41	0
038	Molybdenum (mg / kg (ppm))	0278	1.395	0.2700	1.462	0.4227	0.1394	18	-0.16	038.42	0
038	Molybdenum (mg / kg (ppm))	0407	1.408	0.0297	1.462	0.4227	0.1394	18	-0.13	038.41	0
038	Molybdenum (mg / kg (ppm))	0226	1.510	0.0600	1.462	0.4227	0.1394	18	0.11	038.41	0
038	Molybdenum (mg / kg (ppm))	0693	1.586	0.4310	1.462	0.4227	0.1394	18	0.29	038.42	0
038	Molybdenum (mg / kg (ppm))	0164	1.645	0.0100	1.462	0.4227	0.1394	18	0.43	038.53	0
038	Molybdenum (mg / kg (ppm))	2055	1.650	0.3000	1.462	0.4227	0.1394	18	0.44	038.99	0
038	Molybdenum (mg / kg (ppm))	0297	1.905	0.1500	1.462	0.4227	0.1394	18	1.05	038.43	0
038	Molybdenum (mg / kg (ppm))	0510	2.000	0.0000	1.462	0.4227	0.1394	18	1.27	038.43	0
038	Molybdenum (mg / kg (ppm))	0083	2.050	0.1000	1.462	0.4227	0.1394	18	1.39	038.43	0
038	Molybdenum (mg / kg (ppm))	0265	4.530	0.3400	1.462	0.4227	0.1394	18	7.26	038.42	0
038	Molybdenum (mg / kg (ppm))	0045	1.900	1.400	1.462	0.4227	0.1394	18	1.04	038.42	1
040	Barium (mg / kg (ppm))	0918	17.58	0.5678				2		040.53	0
040	Barium (mg / kg (ppm))	0560	19.86	0.4100				2		040.42	0
042	Chloride (%)	0948	0.9400	0.0000				3		042.00	0
042	Chloride (%)	2200	1.700	0.0000				3		042.00	0
042	Chloride (%)	2199	1.810	0.0200				3		042.00	0
042	Chloride (%)	0297	0.4900	0.0800				3		042.99	1
102	Niacin (mg / kg (ppm))	0227	103.0	0.0000				1		102.01	0
103	Pantothenic Acid (mg / kg (ppm))	0227	10.21	0.7900				1		103.01	0
104	Riboflavin (mg / kg (ppm))	0910	0.7150	0.0300				2		104.03	0
104	Riboflavin (mg / kg (ppm))	0227	3.030	0.1000				2		104.00	0
105	Thiamine (mg / kg (ppm))	0910	4.815	0.5300				2		105.00	0
105	Thiamine (mg / kg (ppm))	0227	6.675	1.610				2		105.01	0
106	Vitamin A (KU / kg)	0563	17.57	0.0539	31.32	5.794	1.905	18	-2.37	106.02	0
106	Vitamin A (KU / kg)	0905	24.87	0.7230	31.32	5.794	1.905	18	-1.11	106.02	0
106	Vitamin A (KU / kg)	2053	25.60	2.800	31.32	5.794	1.905	18	-0.99	106.02	0
106	Vitamin A (KU / kg)	0019	26.79	0.2200	31.32	5.794	1.905	18	-0.78	106.00	0
106	Vitamin A (KU / kg)	0676	27.85	3.100	31.32	5.794	1.905	18	-0.60	106.02	0
106	Vitamin A (KU / kg)	0035	27.89	1.690	31.32	5.794	1.905	18	-0.59	106.02	0
106	Vitamin A (KU / kg)	0512	27.96	2.260	31.32	5.794	1.905	18	-0.58	106.02	0
106	Vitamin A (KU / kg)	0610	30.80	2.200	31.32	5.794	1.905	18	-0.09	106.02	0
106	Vitamin A (KU / kg)	0910	31.05	2.500	31.32	5.794	1.905	18	-0.05	106.02	0
106	Vitamin A (KU / kg)	0098	31.25	2.500	31.32	5.794	1.905	18	-0.01	106.01	0
106	Vitamin A (KU / kg)	0171	32.18	0.3600	31.32	5.794	1.905	18	0.15	106.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
106	Vitamin A (KU / kg)	0148	32.41	0.0200	31.32	5.794	1.905	18	0.19	106.02	0
106	Vitamin A (KU / kg)	0004	32.73	5.350	31.32	5.794	1.905	18	0.24	106.02	0
106	Vitamin A (KU / kg)	0010	36.40	4.000	31.32	5.794	1.905	18	0.88	106.02	0
106	Vitamin A (KU / kg)	2103	36.64	1.390	31.32	5.794	1.905	18	0.92	106.02	0
106	Vitamin A (KU / kg)	0638	36.70	0.0000	31.32	5.794	1.905	18	0.93	106.02	0
106	Vitamin A (KU / kg)	0689	40.75	0.3000	31.32	5.794	1.905	18	1.63	106.02	0
106	Vitamin A (KU / kg)	0675	83.18	4.820	31.32	5.794	1.905	18	8.95	106.02	0
106	Vitamin A (KU / kg)	0038	31.88	9.790	31.32	5.794	1.905	18	0.10	106.02	1
107	Vitamin B12 (µg / kg (ppb))	0227	< 4.4					0		107.00	5
108	Vitamin D3 (KU / kg)	0910	2.025	0.6100				3		108.02	0
108	Vitamin D3 (KU / kg)	0675	4.610	1.380				3		108.02	0
108	Vitamin D3 (KU / kg)	0676	5.575	0.6500				3		108.02	0
108	Vitamin D3 (KU / kg)	0610	< 5000					3		108.02	5
109	Vitamin E (IU / kg)	0905	49.50	9.000	67.00	8.289	2.659	10	-2.11	109.02	0
109	Vitamin E (IU / kg)	2103	59.50	1.000	67.00	8.289	2.659	10	-0.90	109.02	0
109	Vitamin E (IU / kg)	0171	61.50	1.000	67.00	8.289	2.659	10	-0.66	109.99	0
109	Vitamin E (IU / kg)	0910	63.00	1.000	67.00	8.289	2.659	10	-0.48	109.02	0
109	Vitamin E (IU / kg)	0676	67.65	0.5000	67.00	8.289	2.659	10	0.08	109.02	0
109	Vitamin E (IU / kg)	0098	69.85	0.9000	67.00	8.289	2.659	10	0.34	109.02	0
109	Vitamin E (IU / kg)	0227	71.30	3.200	67.00	8.289	2.659	10	0.52	109.02	0
109	Vitamin E (IU / kg)	0148	71.40	7.970	67.00	8.289	2.659	10	0.53	109.02	0
109	Vitamin E (IU / kg)	0638	71.81	0.0200	67.00	8.289	2.659	10	0.58	109.02	0
109	Vitamin E (IU / kg)	0610	164.0	2.000	67.00	8.289	2.659	10	11.70	109.02	0
109	Vitamin E (IU / kg)	0675	221.1	22.18	67.00	8.289	2.659	10	18.59	109.02	1
112	Pyridoxine (µg / g)	0227	4.485	0.0100				1		112.01	0
113	Folic Acid (mg / kg (ppm))	0227	0.8235	0.0550				1		113.01	0
114	Biotin (mg / kg (ppm))	0227	0.2710	0.0400				1		114.01	0
115	Non Protein N (NPN) (%)	0098	1.305	0.0700				1		115.00	0
118	Peroxide value (meq/kg)	2081	9.170	0.9800				1		118.99	0
120	Alanine (%)	2246	0.6163	0.0124	0.7469	0.0346	0.0152	34	-3.77	120.05	0
120	Alanine (%)	0918	0.6645	0.0354	0.7469	0.0346	0.0152	34	-2.38	120.00	0
120	Alanine (%)	0353	0.6950	0.0100	0.7469	0.0346	0.0152	34	-1.50	120.00	0
120	Alanine (%)	2256	0.7040	0.0020	0.7469	0.0346	0.0152	34	-1.24	120.00	0
120	Alanine (%)	2144	0.7050	0.0500	0.7469	0.0346	0.0152	34	-1.21	120.99	0
120	Alanine (%)	0675	0.7100	0.0200	0.7469	0.0346	0.0152	34	-1.07	120.00	0
120	Alanine (%)	0910	0.7200	0.0400	0.7469	0.0346	0.0152	34	-0.78	120.00	0
120	Alanine (%)	0676	0.7250	0.0100	0.7469	0.0346	0.0152	34	-0.63	120.05	0
120	Alanine (%)	0226	0.7254	0.0687	0.7469	0.0346	0.0152	34	-0.62	120.00	0
120	Alanine (%)	0682	0.7280	0.0000	0.7469	0.0346	0.0152	34	-0.55	120.00	0
120	Alanine (%)	0610	0.7300	0.0000	0.7469	0.0346	0.0152	34	-0.49	120.05	0
120	Alanine (%)	0407	0.7350	0.0260	0.7469	0.0346	0.0152	34	-0.34	120.05	0
120	Alanine (%)	0504	0.7400	0.0200	0.7469	0.0346	0.0152	34	-0.20	120.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
120	Alanine (%)	0148	0.7400	0.0200	0.7469	0.0346	0.0152	34	-0.20	120.05	0
120	Alanine (%)	0619	0.7420	0.0020	0.7469	0.0346	0.0152	34	-0.14	120.00	0
120	Alanine (%)	0227	0.7450	0.0100	0.7469	0.0346	0.0152	34	-0.05	120.00	0
120	Alanine (%)	0868	0.7455	0.0050	0.7469	0.0346	0.0152	34	-0.04	120.00	0
120	Alanine (%)	0968	0.7460	0.0080	0.7469	0.0346	0.0152	34	-0.03	120.00	0
120	Alanine (%)	0878	0.7475	0.0010	0.7469	0.0346	0.0152	34	0.02	120.00	0
120	Alanine (%)	0652	0.7550	0.0100	0.7469	0.0346	0.0152	34	0.23	120.00	0
120	Alanine (%)	0872	0.7550	0.0080	0.7469	0.0346	0.0152	34	0.23	120.00	0
120	Alanine (%)	0571	0.7565	0.0170	0.7469	0.0346	0.0152	34	0.28	120.00	0
120	Alanine (%)	2059	0.7610	0.0040	0.7469	0.0346	0.0152	34	0.41	120.00	0
120	Alanine (%)	0684	0.7650	0.0100	0.7469	0.0346	0.0152	34	0.52	120.00	0
120	Alanine (%)	0242	0.7650	0.0500	0.7469	0.0346	0.0152	34	0.52	120.05	0
120	Alanine (%)	0644	0.7690	0.0040	0.7469	0.0346	0.0152	34	0.64	120.00	0
120	Alanine (%)	0098	0.7745	0.0010	0.7469	0.0346	0.0152	34	0.80	120.02	0
120	Alanine (%)	2196	0.7750	0.0000	0.7469	0.0346	0.0152	34	0.81	120.05	0
120	Alanine (%)	0859	0.7791	0.0175	0.7469	0.0346	0.0152	34	0.93	120.00	0
120	Alanine (%)	0941	0.7795	0.0130	0.7469	0.0346	0.0152	34	0.94	120.00	0
120	Alanine (%)	0939	0.7900	0.0000	0.7469	0.0346	0.0152	34	1.25	120.00	0
120	Alanine (%)	2188	0.8050	0.0100	0.7469	0.0346	0.0152	34	1.68	120.05	0
120	Alanine (%)	0870	0.8394	0.0234	0.7469	0.0346	0.0152	34	2.67	120.00	0
120	Alanine (%)	2058	1.465	0.0100	0.7469	0.0346	0.0152	34	20.75	120.99	0
120	Alanine (%)	0626	759.5	19.00	0.7469	0.0346	0.0152	34	21921.29	120.05	2
121	Arginine (%)	0407	0.8730	0.0720	1.069	0.0619	0.0231	33	-3.16	121.05	0
121	Arginine (%)	0918	0.9148	0.0413	1.069	0.0619	0.0231	33	-2.49	121.00	0
121	Arginine (%)	0675	0.9400	0.0200	1.069	0.0619	0.0231	33	-2.08	121.00	0
121	Arginine (%)	0682	0.9700	0.0000	1.069	0.0619	0.0231	33	-1.59	121.00	0
121	Arginine (%)	0626	0.9830	0.0040	1.069	0.0619	0.0231	33	-1.38	121.05	0
121	Arginine (%)	2256	1.013	0.0030	1.069	0.0619	0.0231	33	-0.91	121.00	0
121	Arginine (%)	0619	1.016	0.0680	1.069	0.0619	0.0231	33	-0.85	121.00	0
121	Arginine (%)	0910	1.025	0.0500	1.069	0.0619	0.0231	33	-0.71	121.00	0
121	Arginine (%)	0504	1.035	0.0300	1.069	0.0619	0.0231	33	-0.54	121.00	0
121	Arginine (%)	0098	1.045	0.0290	1.069	0.0619	0.0231	33	-0.39	121.02	0
121	Arginine (%)	0227	1.045	0.0100	1.069	0.0619	0.0231	33	-0.38	121.00	0
121	Arginine (%)	0968	1.045	0.0060	1.069	0.0619	0.0231	33	-0.38	121.00	0
121	Arginine (%)	0242	1.045	0.0100	1.069	0.0619	0.0231	33	-0.38	121.05	0
121	Arginine (%)	0868	1.060	0.0100	1.069	0.0619	0.0231	33	-0.14	121.00	0
121	Arginine (%)	0684	1.075	0.0500	1.069	0.0619	0.0231	33	0.10	121.00	0
121	Arginine (%)	0676	1.075	0.0100	1.069	0.0619	0.0231	33	0.10	121.05	0
121	Arginine (%)	0226	1.085	0.0196	1.069	0.0619	0.0231	33	0.26	121.00	0
121	Arginine (%)	0939	1.090	0.0000	1.069	0.0619	0.0231	33	0.34	121.00	0
121	Arginine (%)	2196	1.091	0.0000	1.069	0.0619	0.0231	33	0.36	121.05	0
121	Arginine (%)	0872	1.091	0.0180	1.069	0.0619	0.0231	33	0.36	121.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
121	Arginine (%)	0644	1.094	0.0020	1.069	0.0619	0.0231	33	0.41	121.00	0
121	Arginine (%)	0610	1.095	0.0100	1.069	0.0619	0.0231	33	0.43	121.05	0
121	Arginine (%)	0859	1.097	0.0279	1.069	0.0619	0.0231	33	0.46	121.00	0
121	Arginine (%)	0652	1.100	0.0200	1.069	0.0619	0.0231	33	0.51	121.00	0
121	Arginine (%)	0148	1.100	0.0600	1.069	0.0619	0.0231	33	0.51	121.05	0
121	Arginine (%)	0878	1.106	0.0170	1.069	0.0619	0.0231	33	0.60	121.00	0
121	Arginine (%)	0571	1.106	0.0040	1.069	0.0619	0.0231	33	0.60	121.00	0
121	Arginine (%)	2059	1.113	0.0030	1.069	0.0619	0.0231	33	0.71	121.00	0
121	Arginine (%)	0941	1.129	0.0130	1.069	0.0619	0.0231	33	0.97	121.00	0
121	Arginine (%)	2188	1.130	0.0600	1.069	0.0619	0.0231	33	0.99	121.05	0
121	Arginine (%)	0870	1.153	0.0435	1.069	0.0619	0.0231	33	1.36	121.00	0
121	Arginine (%)	2058	1.170	0.0000	1.069	0.0619	0.0231	33	1.64	121.99	0
121	Arginine (%)	0353	1.235	0.0500	1.069	0.0619	0.0231	33	2.69	121.00	0
122	Aspartic (%)	2246	0.8998	0.0459	1.155	0.0605	0.0261	35	-4.22	122.05	0
122	Aspartic (%)	0675	1.010	0.0200	1.155	0.0605	0.0261	35	-2.39	122.00	0
122	Aspartic (%)	0353	1.030	0.0200	1.155	0.0605	0.0261	35	-2.06	122.00	0
122	Aspartic (%)	0682	1.049	0.0000	1.155	0.0605	0.0261	35	-1.75	122.00	0
122	Aspartic (%)	0676	1.065	0.0500	1.155	0.0605	0.0261	35	-1.49	122.05	0
122	Aspartic (%)	2256	1.080	0.0030	1.155	0.0605	0.0261	35	-1.25	122.00	0
122	Aspartic (%)	0626	1.089	0.0130	1.155	0.0605	0.0261	35	-1.10	122.05	0
122	Aspartic (%)	0918	1.107	0.0481	1.155	0.0605	0.0261	35	-0.79	122.00	0
122	Aspartic (%)	0407	1.114	0.0560	1.155	0.0605	0.0261	35	-0.68	122.05	0
122	Aspartic (%)	0226	1.124	0.0255	1.155	0.0605	0.0261	35	-0.51	122.00	0
122	Aspartic (%)	0227	1.145	0.0100	1.155	0.0605	0.0261	35	-0.16	122.00	0
122	Aspartic (%)	0910	1.145	0.0500	1.155	0.0605	0.0261	35	-0.16	122.00	0
122	Aspartic (%)	0968	1.146	0.0520	1.155	0.0605	0.0261	35	-0.15	122.00	0
122	Aspartic (%)	0504	1.150	0.0200	1.155	0.0605	0.0261	35	-0.08	122.00	0
122	Aspartic (%)	0644	1.157	0.0100	1.155	0.0605	0.0261	35	0.04	122.00	0
122	Aspartic (%)	0872	1.159	0.0320	1.155	0.0605	0.0261	35	0.07	122.00	0
122	Aspartic (%)	0571	1.160	0.0010	1.155	0.0605	0.0261	35	0.08	122.00	0
122	Aspartic (%)	2059	1.164	0.0000	1.155	0.0605	0.0261	35	0.15	122.00	0
122	Aspartic (%)	0610	1.165	0.0100	1.155	0.0605	0.0261	35	0.17	122.05	0
122	Aspartic (%)	2144	1.175	0.0900	1.155	0.0605	0.0261	35	0.33	122.99	0
122	Aspartic (%)	0859	1.175	0.0271	1.155	0.0605	0.0261	35	0.34	122.00	0
122	Aspartic (%)	0652	1.180	0.0000	1.155	0.0605	0.0261	35	0.42	122.00	0
122	Aspartic (%)	0878	1.183	0.0190	1.155	0.0605	0.0261	35	0.46	122.00	0
122	Aspartic (%)	0684	1.185	0.0300	1.155	0.0605	0.0261	35	0.50	122.00	0
122	Aspartic (%)	2188	1.185	0.1100	1.155	0.0605	0.0261	35	0.50	122.05	0
122	Aspartic (%)	0098	1.186	0.0220	1.155	0.0605	0.0261	35	0.51	122.02	0
122	Aspartic (%)	0941	1.190	0.0010	1.155	0.0605	0.0261	35	0.57	122.00	0
122	Aspartic (%)	0939	1.190	0.0000	1.155	0.0605	0.0261	35	0.58	122.00	0
122	Aspartic (%)	0868	1.195	0.0010	1.155	0.0605	0.0261	35	0.66	122.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
122	Aspartic (%)	0148	1.205	0.0100	1.155	0.0605	0.0261	35	0.83	122.05	0
122	Aspartic (%)	0870	1.212	0.0392	1.155	0.0605	0.0261	35	0.94	122.00	0
122	Aspartic (%)	0242	1.220	0.0800	1.155	0.0605	0.0261	35	1.08	122.05	0
122	Aspartic (%)	0619	1.224	0.0080	1.155	0.0605	0.0261	35	1.14	122.00	0
122	Aspartic (%)	2196	1.260	0.0000	1.155	0.0605	0.0261	35	1.74	122.05	0
122	Aspartic (%)	2058	1.465	0.0100	1.155	0.0605	0.0261	35	5.13	122.99	0
124	Cysteine/Cystine (%)	2246	0.0508	0.0001	0.2995	0.0276	0.0093	33	-9.01	124.05	0
124	Cysteine/Cystine (%)	2144	0.2150	0.0300	0.2995	0.0276	0.0093	33	-3.06	124.99	0
124	Cysteine/Cystine (%)	0353	0.2350	0.0100	0.2995	0.0276	0.0093	33	-2.34	124.00	0
124	Cysteine/Cystine (%)	0148	0.2550	0.0100	0.2995	0.0276	0.0093	33	-1.61	124.05	0
124	Cysteine/Cystine (%)	0407	0.2675	0.0290	0.2995	0.0276	0.0093	33	-1.16	124.05	0
124	Cysteine/Cystine (%)	0870	0.2700	0.0000	0.2995	0.0276	0.0093	33	-1.07	124.00	0
124	Cysteine/Cystine (%)	0227	0.2750	0.0100	0.2995	0.0276	0.0093	33	-0.89	124.00	0
124	Cysteine/Cystine (%)	2188	0.2850	0.0100	0.2995	0.0276	0.0093	33	-0.52	124.05	0
124	Cysteine/Cystine (%)	0644	0.2905	0.0050	0.2995	0.0276	0.0093	33	-0.33	124.00	0
124	Cysteine/Cystine (%)	0872	0.2920	0.0080	0.2995	0.0276	0.0093	33	-0.27	124.00	0
124	Cysteine/Cystine (%)	0968	0.2930	0.0020	0.2995	0.0276	0.0093	33	-0.23	124.00	0
124	Cysteine/Cystine (%)	0652	0.2950	0.0100	0.2995	0.0276	0.0093	33	-0.16	124.00	0
124	Cysteine/Cystine (%)	2196	0.2950	0.0000	0.2995	0.0276	0.0093	33	-0.16	124.05	0
124	Cysteine/Cystine (%)	0571	0.2975	0.0030	0.2995	0.0276	0.0093	33	-0.07	124.00	0
124	Cysteine/Cystine (%)	0682	0.2990	0.0000	0.2995	0.0276	0.0093	33	-0.02	124.00	0
124	Cysteine/Cystine (%)	0939	0.3000	0.0000	0.2995	0.0276	0.0093	33	0.02	124.00	0
124	Cysteine/Cystine (%)	0676	0.3000	0.0000	0.2995	0.0276	0.0093	33	0.02	124.05	0
124	Cysteine/Cystine (%)	0859	0.3001	0.0003	0.2995	0.0276	0.0093	33	0.02	124.00	0
124	Cysteine/Cystine (%)	0504	0.3050	0.0100	0.2995	0.0276	0.0093	33	0.20	124.00	0
124	Cysteine/Cystine (%)	0910	0.3050	0.0100	0.2995	0.0276	0.0093	33	0.20	124.00	0
124	Cysteine/Cystine (%)	2059	0.3055	0.0030	0.2995	0.0276	0.0093	33	0.22	124.00	0
124	Cysteine/Cystine (%)	0941	0.3065	0.0050	0.2995	0.0276	0.0093	33	0.25	124.00	0
124	Cysteine/Cystine (%)	0878	0.3070	0.0020	0.2995	0.0276	0.0093	33	0.27	124.00	0
124	Cysteine/Cystine (%)	0684	0.3100	0.0000	0.2995	0.0276	0.0093	33	0.38	124.00	0
124	Cysteine/Cystine (%)	0160	0.3100	0.0400	0.2995	0.0276	0.0093	33	0.38	124.99	0
124	Cysteine/Cystine (%)	0098	0.3130	0.0080	0.2995	0.0276	0.0093	33	0.49	124.02	0
124	Cysteine/Cystine (%)	0226	0.3142	0.0028	0.2995	0.0276	0.0093	33	0.53	124.00	0
124	Cysteine/Cystine (%)	0868	0.3170	0.0060	0.2995	0.0276	0.0093	33	0.63	124.00	0
124	Cysteine/Cystine (%)	0610	0.3350	0.0100	0.2995	0.0276	0.0093	33	1.29	124.05	0
124	Cysteine/Cystine (%)	0918	0.3400	0.0400	0.2995	0.0276	0.0093	33	1.47	124.00	0
124	Cysteine/Cystine (%)	0242	0.3450	0.0100	0.2995	0.0276	0.0093	33	1.65	124.05	0
124	Cysteine/Cystine (%)	0619	0.3580	0.0140	0.2995	0.0276	0.0093	33	2.12	124.00	0
124	Cysteine/Cystine (%)	0675	0.3800	0.0200	0.2995	0.0276	0.0093	33	2.92	124.00	0
125	Glutamic (%)	0675	2.590	0.0200	2.905	0.1420	0.0571	34	-2.22	125.00	0
125	Glutamic (%)	2246	2.633	0.0713	2.905	0.1420	0.0571	34	-1.91	125.05	0
125	Glutamic (%)	0676	2.720	0.0000	2.905	0.1420	0.0571	34	-1.30	125.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
125	Glutamic (%)	2256	2.751	0.0020	2.905	0.1420	0.0571	34	-1.08	125.00	0
125	Glutamic (%)	0682	2.761	0.0000	2.905	0.1420	0.0571	34	-1.01	125.00	0
125	Glutamic (%)	0918	2.774	0.0540	2.905	0.1420	0.0571	34	-0.92	125.00	0
125	Glutamic (%)	0353	2.775	0.0300	2.905	0.1420	0.0571	34	-0.91	125.00	0
125	Glutamic (%)	0941	2.782	0.0670	2.905	0.1420	0.0571	34	-0.87	125.00	0
125	Glutamic (%)	0504	2.815	0.0900	2.905	0.1420	0.0571	34	-0.63	125.00	0
125	Glutamic (%)	2196	2.845	0.0000	2.905	0.1420	0.0571	34	-0.42	125.05	0
125	Glutamic (%)	0227	2.850	0.0000	2.905	0.1420	0.0571	34	-0.38	125.00	0
125	Glutamic (%)	2059	2.854	0.0080	2.905	0.1420	0.0571	34	-0.36	125.00	0
125	Glutamic (%)	0610	2.855	0.0100	2.905	0.1420	0.0571	34	-0.35	125.05	0
125	Glutamic (%)	0571	2.859	0.0420	2.905	0.1420	0.0571	34	-0.32	125.00	0
125	Glutamic (%)	0868	2.873	0.0170	2.905	0.1420	0.0571	34	-0.23	125.00	0
125	Glutamic (%)	0098	2.894	0.0600	2.905	0.1420	0.0571	34	-0.07	125.02	0
125	Glutamic (%)	0859	2.899	0.0184	2.905	0.1420	0.0571	34	-0.04	125.00	0
125	Glutamic (%)	0968	2.908	0.0190	2.905	0.1420	0.0571	34	0.02	125.00	0
125	Glutamic (%)	0872	2.908	0.0690	2.905	0.1420	0.0571	34	0.02	125.00	0
125	Glutamic (%)	0407	2.912	0.1600	2.905	0.1420	0.0571	34	0.05	125.05	0
125	Glutamic (%)	0910	2.915	0.0700	2.905	0.1420	0.0571	34	0.07	125.00	0
125	Glutamic (%)	0652	2.940	0.0400	2.905	0.1420	0.0571	34	0.25	125.00	0
125	Glutamic (%)	0626	2.957	0.0310	2.905	0.1420	0.0571	34	0.37	125.05	0
125	Glutamic (%)	0644	2.959	0.0040	2.905	0.1420	0.0571	34	0.38	125.00	0
125	Glutamic (%)	0878	2.962	0.0150	2.905	0.1420	0.0571	34	0.40	125.00	0
125	Glutamic (%)	0242	2.965	0.2300	2.905	0.1420	0.0571	34	0.43	125.05	0
125	Glutamic (%)	0684	2.980	0.2200	2.905	0.1420	0.0571	34	0.53	125.00	0
125	Glutamic (%)	0619	3.055	0.0100	2.905	0.1420	0.0571	34	1.06	125.00	0
125	Glutamic (%)	0939	3.070	0.0000	2.905	0.1420	0.0571	34	1.16	125.00	0
125	Glutamic (%)	2058	3.080	0.0400	2.905	0.1420	0.0571	34	1.24	125.99	0
125	Glutamic (%)	0226	3.105	0.1031	2.905	0.1420	0.0571	34	1.41	125.00	0
125	Glutamic (%)	0148	3.180	0.0200	2.905	0.1420	0.0571	34	1.94	125.05	0
125	Glutamic (%)	0870	3.291	0.1704	2.905	0.1420	0.0571	34	2.72	125.00	0
125	Glutamic (%)	2144	3.785	0.2500	2.905	0.1420	0.0571	34	6.20	125.99	0
125	Glutamic (%)	2188	2.610	0.4200	2.905	0.1420	0.0571	34	-2.07	125.05	1
126	Glycine (%)	2144	0.6650	0.0500	0.7613	0.0383	0.0192	35	-2.51	126.99	0
126	Glycine (%)	0675	0.6800	0.0200	0.7613	0.0383	0.0192	35	-2.12	126.00	0
126	Glycine (%)	0682	0.7010	0.0000	0.7613	0.0383	0.0192	35	-1.57	126.00	0
126	Glycine (%)	0918	0.7067	0.0517	0.7613	0.0383	0.0192	35	-1.43	126.00	0
126	Glycine (%)	2246	0.7167	0.0079	0.7613	0.0383	0.0192	35	-1.16	126.05	0
126	Glycine (%)	2256	0.7185	0.0050	0.7613	0.0383	0.0192	35	-1.12	126.00	0
126	Glycine (%)	0353	0.7250	0.0100	0.7613	0.0383	0.0192	35	-0.95	126.00	0
126	Glycine (%)	0504	0.7350	0.0300	0.7613	0.0383	0.0192	35	-0.69	126.00	0
126	Glycine (%)	0407	0.7380	0.0200	0.7613	0.0383	0.0192	35	-0.61	126.05	0
126	Glycine (%)	0910	0.7400	0.0200	0.7613	0.0383	0.0192	35	-0.56	126.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
126	Glycine (%)	0676	0.7400	0.0400	0.7613	0.0383	0.0192	35	-0.56	126.05	0
126	Glycine (%)	0242	0.7450	0.0500	0.7613	0.0383	0.0192	35	-0.43	126.05	0
126	Glycine (%)	0968	0.7595	0.0130	0.7613	0.0383	0.0192	35	-0.05	126.00	0
126	Glycine (%)	0227	0.7600	0.0000	0.7613	0.0383	0.0192	35	-0.03	126.00	0
126	Glycine (%)	0652	0.7600	0.0200	0.7613	0.0383	0.0192	35	-0.03	126.00	0
126	Glycine (%)	0868	0.7600	0.0040	0.7613	0.0383	0.0192	35	-0.03	126.00	0
126	Glycine (%)	0571	0.7630	0.0100	0.7613	0.0383	0.0192	35	0.04	126.00	0
126	Glycine (%)	0148	0.7650	0.0100	0.7613	0.0383	0.0192	35	0.10	126.05	0
126	Glycine (%)	0610	0.7650	0.0100	0.7613	0.0383	0.0192	35	0.10	126.05	0
126	Glycine (%)	0872	0.7655	0.0190	0.7613	0.0383	0.0192	35	0.11	126.00	0
126	Glycine (%)	2059	0.7690	0.0040	0.7613	0.0383	0.0192	35	0.20	126.00	0
126	Glycine (%)	0684	0.7700	0.0200	0.7613	0.0383	0.0192	35	0.23	126.00	0
126	Glycine (%)	0098	0.7705	0.0030	0.7613	0.0383	0.0192	35	0.24	126.02	0
126	Glycine (%)	0644	0.7745	0.0010	0.7613	0.0383	0.0192	35	0.34	126.00	0
126	Glycine (%)	0878	0.7745	0.0030	0.7613	0.0383	0.0192	35	0.34	126.00	0
126	Glycine (%)	0619	0.7785	0.0050	0.7613	0.0383	0.0192	35	0.45	126.00	0
126	Glycine (%)	0859	0.7840	0.0137	0.7613	0.0383	0.0192	35	0.59	126.00	0
126	Glycine (%)	0941	0.7930	0.0060	0.7613	0.0383	0.0192	35	0.83	126.00	0
126	Glycine (%)	0939	0.7950	0.0100	0.7613	0.0383	0.0192	35	0.88	126.00	0
126	Glycine (%)	0626	0.7975	0.0250	0.7613	0.0383	0.0192	35	0.94	126.05	0
126	Glycine (%)	2188	0.8050	0.0700	0.7613	0.0383	0.0192	35	1.14	126.05	0
126	Glycine (%)	2196	0.8110	0.0000	0.7613	0.0383	0.0192	35	1.30	126.05	0
126	Glycine (%)	0226	0.8116	0.0813	0.7613	0.0383	0.0192	35	1.31	126.00	0
126	Glycine (%)	0870	0.8321	0.0398	0.7613	0.0383	0.0192	35	1.85	126.00	0
126	Glycine (%)	2058	0.9100	0.0000	0.7613	0.0383	0.0192	35	3.88	126.99	0
127	Histidine (%)	0353	0.3550	0.0100	0.4071	0.0207	0.0112	34	-2.51	127.00	0
127	Histidine (%)	2256	0.3655	0.0010	0.4071	0.0207	0.0112	34	-2.01	127.00	0
127	Histidine (%)	0676	0.3750	0.0100	0.4071	0.0207	0.0112	34	-1.55	127.05	0
127	Histidine (%)	0675	0.3800	0.0200	0.4071	0.0207	0.0112	34	-1.31	127.00	0
127	Histidine (%)	0918	0.3831	0.0035	0.4071	0.0207	0.0112	34	-1.16	127.00	0
127	Histidine (%)	2188	0.3850	0.0300	0.4071	0.0207	0.0112	34	-1.07	127.05	0
127	Histidine (%)	0868	0.3920	0.0040	0.4071	0.0207	0.0112	34	-0.73	127.00	0
127	Histidine (%)	0682	0.3940	0.0000	0.4071	0.0207	0.0112	34	-0.63	127.00	0
127	Histidine (%)	0242	0.3950	0.0300	0.4071	0.0207	0.0112	34	-0.59	127.05	0
127	Histidine (%)	0968	0.3965	0.0130	0.4071	0.0207	0.0112	34	-0.51	127.00	0
127	Histidine (%)	2246	0.3999	0.0069	0.4071	0.0207	0.0112	34	-0.35	127.05	0
127	Histidine (%)	0098	0.4000	0.0040	0.4071	0.0207	0.0112	34	-0.34	127.02	0
127	Histidine (%)	0571	0.4045	0.0030	0.4071	0.0207	0.0112	34	-0.13	127.00	0
127	Histidine (%)	0227	0.4050	0.0100	0.4071	0.0207	0.0112	34	-0.10	127.00	0
127	Histidine (%)	0504	0.4050	0.0100	0.4071	0.0207	0.0112	34	-0.10	127.00	0
127	Histidine (%)	0910	0.4050	0.0100	0.4071	0.0207	0.0112	34	-0.10	127.00	0
127	Histidine (%)	0610	0.4050	0.0100	0.4071	0.0207	0.0112	34	-0.10	127.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
127	Histidine (%)	0626	0.4080	0.0200	0.4071	0.0207	0.0112	34	0.04	127.05	0
127	Histidine (%)	0226	0.4119	0.0277	0.4071	0.0207	0.0112	34	0.23	127.00	0
127	Histidine (%)	0878	0.4120	0.0080	0.4071	0.0207	0.0112	34	0.23	127.00	0
127	Histidine (%)	2059	0.4125	0.0010	0.4071	0.0207	0.0112	34	0.26	127.00	0
127	Histidine (%)	0859	0.4128	0.0085	0.4071	0.0207	0.0112	34	0.27	127.00	0
127	Histidine (%)	0872	0.4130	0.0080	0.4071	0.0207	0.0112	34	0.28	127.00	0
127	Histidine (%)	0644	0.4145	0.0010	0.4071	0.0207	0.0112	34	0.35	127.00	0
127	Histidine (%)	0684	0.4150	0.0100	0.4071	0.0207	0.0112	34	0.38	127.00	0
127	Histidine (%)	0870	0.4168	0.0181	0.4071	0.0207	0.0112	34	0.46	127.00	0
127	Histidine (%)	0941	0.4195	0.0070	0.4071	0.0207	0.0112	34	0.60	127.00	0
127	Histidine (%)	0939	0.4200	0.0000	0.4071	0.0207	0.0112	34	0.62	127.00	0
127	Histidine (%)	0148	0.4200	0.0200	0.4071	0.0207	0.0112	34	0.62	127.05	0
127	Histidine (%)	0407	0.4360	0.0340	0.4071	0.0207	0.0112	34	1.39	127.05	0
127	Histidine (%)	0652	0.4400	0.0000	0.4071	0.0207	0.0112	34	1.58	127.00	0
127	Histidine (%)	2058	0.4400	0.0200	0.4071	0.0207	0.0112	34	1.58	127.99	0
127	Histidine (%)	2196	0.4700	0.0000	0.4071	0.0207	0.0112	34	3.03	127.05	0
127	Histidine (%)	0619	0.5175	0.0230	0.4071	0.0207	0.0112	34	5.32	127.00	0
127	Histidine (%)	2144	0.4350	0.1100	0.4071	0.0207	0.0112	34	1.34	127.99	1
128	Isoleucine (%)	0619	0.3705	0.0630	0.5008	0.0402	0.0172	35	-3.24	128.00	0
128	Isoleucine (%)	0353	0.3850	0.0100	0.5008	0.0402	0.0172	35	-2.88	128.00	0
128	Isoleucine (%)	0675	0.3900	0.0200	0.5008	0.0402	0.0172	35	-2.75	128.00	0
128	Isoleucine (%)	0407	0.4305	0.0250	0.5008	0.0402	0.0172	35	-1.75	128.05	0
128	Isoleucine (%)	0918	0.4407	0.0245	0.5008	0.0402	0.0172	35	-1.50	128.00	0
128	Isoleucine (%)	2256	0.4505	0.0030	0.5008	0.0402	0.0172	35	-1.25	128.00	0
128	Isoleucine (%)	0684	0.4650	0.0500	0.5008	0.0402	0.0172	35	-0.89	128.00	0
128	Isoleucine (%)	2246	0.4680	0.0032	0.5008	0.0402	0.0172	35	-0.82	128.05	0
128	Isoleucine (%)	0676	0.4750	0.0100	0.5008	0.0402	0.0172	35	-0.64	128.05	0
128	Isoleucine (%)	0868	0.4775	0.0050	0.5008	0.0402	0.0172	35	-0.58	128.00	0
128	Isoleucine (%)	0682	0.4830	0.0000	0.5008	0.0402	0.0172	35	-0.44	128.00	0
128	Isoleucine (%)	0968	0.4905	0.0150	0.5008	0.0402	0.0172	35	-0.26	128.00	0
128	Isoleucine (%)	0226	0.4966	0.0653	0.5008	0.0402	0.0172	35	-0.11	128.00	0
128	Isoleucine (%)	0242	0.5000	0.0200	0.5008	0.0402	0.0172	35	-0.02	128.05	0
128	Isoleucine (%)	0872	0.5035	0.0130	0.5008	0.0402	0.0172	35	0.07	128.00	0
128	Isoleucine (%)	0859	0.5043	0.0106	0.5008	0.0402	0.0172	35	0.09	128.00	0
128	Isoleucine (%)	0878	0.5075	0.0030	0.5008	0.0402	0.0172	35	0.17	128.00	0
128	Isoleucine (%)	0098	0.5085	0.0090	0.5008	0.0402	0.0172	35	0.19	128.02	0
128	Isoleucine (%)	0227	0.5100	0.0400	0.5008	0.0402	0.0172	35	0.23	128.00	0
128	Isoleucine (%)	2188	0.5100	0.0200	0.5008	0.0402	0.0172	35	0.23	128.05	0
128	Isoleucine (%)	0571	0.5130	0.0080	0.5008	0.0402	0.0172	35	0.30	128.00	0
128	Isoleucine (%)	0652	0.5150	0.0100	0.5008	0.0402	0.0172	35	0.35	128.00	0
128	Isoleucine (%)	2196	0.5150	0.0000	0.5008	0.0402	0.0172	35	0.35	128.05	0
128	Isoleucine (%)	0939	0.5200	0.0200	0.5008	0.0402	0.0172	35	0.48	128.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
128	Isoleucine (%)	0610	0.5200	0.0000	0.5008	0.0402	0.0172	35	0.48	128.05	0
128	Isoleucine (%)	0626	0.5245	0.0030	0.5008	0.0402	0.0172	35	0.59	128.05	0
128	Isoleucine (%)	0910	0.5250	0.0300	0.5008	0.0402	0.0172	35	0.60	128.00	0
128	Isoleucine (%)	2144	0.5250	0.0100	0.5008	0.0402	0.0172	35	0.60	128.99	0
128	Isoleucine (%)	2059	0.5315	0.0030	0.5008	0.0402	0.0172	35	0.76	128.00	0
128	Isoleucine (%)	2058	0.5350	0.0500	0.5008	0.0402	0.0172	35	0.85	128.99	0
128	Isoleucine (%)	0941	0.5395	0.0070	0.5008	0.0402	0.0172	35	0.96	128.00	0
128	Isoleucine (%)	0504	0.5450	0.0100	0.5008	0.0402	0.0172	35	1.10	128.00	0
128	Isoleucine (%)	0644	0.5460	0.0020	0.5008	0.0402	0.0172	35	1.12	128.00	0
128	Isoleucine (%)	0148	0.5600	0.0000	0.5008	0.0402	0.0172	35	1.47	128.05	0
128	Isoleucine (%)	0870	0.5784	0.0394	0.5008	0.0402	0.0172	35	1.93	128.00	0
129	Leucine (%)	2246	0.9409	0.0049	1.069	0.0417	0.0215	35	-3.07	129.05	0
129	Leucine (%)	0675	0.9600	0.0200	1.069	0.0417	0.0215	35	-2.61	129.00	0
129	Leucine (%)	0353	0.9750	0.0100	1.069	0.0417	0.0215	35	-2.25	129.00	0
129	Leucine (%)	0918	1.010	0.0386	1.069	0.0417	0.0215	35	-1.41	129.00	0
129	Leucine (%)	0407	1.012	0.0490	1.069	0.0417	0.0215	35	-1.38	129.05	0
129	Leucine (%)	0682	1.015	0.0000	1.069	0.0417	0.0215	35	-1.29	129.00	0
129	Leucine (%)	0619	1.017	0.0460	1.069	0.0417	0.0215	35	-1.24	129.00	0
129	Leucine (%)	0676	1.020	0.0200	1.069	0.0417	0.0215	35	-1.17	129.05	0
129	Leucine (%)	2256	1.055	0.0010	1.069	0.0417	0.0215	35	-0.34	129.00	0
129	Leucine (%)	2196	1.055	0.0000	1.069	0.0417	0.0215	35	-0.33	129.05	0
129	Leucine (%)	0910	1.060	0.0200	1.069	0.0417	0.0215	35	-0.21	129.00	0
129	Leucine (%)	0652	1.065	0.0100	1.069	0.0417	0.0215	35	-0.09	129.00	0
129	Leucine (%)	0227	1.070	0.0400	1.069	0.0417	0.0215	35	0.03	129.00	0
129	Leucine (%)	0684	1.070	0.0400	1.069	0.0417	0.0215	35	0.03	129.00	0
129	Leucine (%)	0571	1.072	0.0050	1.069	0.0417	0.0215	35	0.06	129.00	0
129	Leucine (%)	0868	1.074	0.0050	1.069	0.0417	0.0215	35	0.11	129.00	0
129	Leucine (%)	0098	1.075	0.0100	1.069	0.0417	0.0215	35	0.15	129.02	0
129	Leucine (%)	0242	1.075	0.0700	1.069	0.0417	0.0215	35	0.15	129.05	0
129	Leucine (%)	0878	1.076	0.0060	1.069	0.0417	0.0215	35	0.17	129.00	0
129	Leucine (%)	0872	1.077	0.0160	1.069	0.0417	0.0215	35	0.19	129.00	0
129	Leucine (%)	0968	1.079	0.0270	1.069	0.0417	0.0215	35	0.23	129.00	0
129	Leucine (%)	2059	1.082	0.0020	1.069	0.0417	0.0215	35	0.31	129.00	0
129	Leucine (%)	0504	1.085	0.0100	1.069	0.0417	0.0215	35	0.39	129.00	0
129	Leucine (%)	0148	1.085	0.0100	1.069	0.0417	0.0215	35	0.39	129.05	0
129	Leucine (%)	0610	1.085	0.0100	1.069	0.0417	0.0215	35	0.39	129.05	0
129	Leucine (%)	2058	1.090	0.0800	1.069	0.0417	0.0215	35	0.51	129.99	0
129	Leucine (%)	2188	1.090	0.0200	1.069	0.0417	0.0215	35	0.51	129.05	0
129	Leucine (%)	0859	1.091	0.0245	1.069	0.0417	0.0215	35	0.54	129.00	0
129	Leucine (%)	0626	1.098	0.0140	1.069	0.0417	0.0215	35	0.70	129.05	0
129	Leucine (%)	0644	1.105	0.0050	1.069	0.0417	0.0215	35	0.85	129.00	0
129	Leucine (%)	0941	1.109	0.0060	1.069	0.0417	0.0215	35	0.96	129.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
129	Leucine (%)	0939	1.110	0.0000	1.069	0.0417	0.0215	35	0.98	129.00	0
129	Leucine (%)	2144	1.125	0.0100	1.069	0.0417	0.0215	35	1.34	129.99	0
129	Leucine (%)	0226	1.134	0.0796	1.069	0.0417	0.0215	35	1.56	129.00	0
129	Leucine (%)	0870	1.177	0.0425	1.069	0.0417	0.0215	35	2.60	129.00	0
130	L-Lysine (%)	2246	0.4961	0.0288	0.6376	0.0391	0.0140	37	-3.62	130.05	0
130	L-Lysine (%)	2053	0.5350	0.0580	0.6376	0.0391	0.0140	37	-2.62	130.99	0
130	L-Lysine (%)	0353	0.5600	0.0000	0.6376	0.0391	0.0140	37	-1.98	130.00	0
130	L-Lysine (%)	0918	0.5695	0.0149	0.6376	0.0391	0.0140	37	-1.74	130.00	0
130	L-Lysine (%)	0675	0.5700	0.0200	0.6376	0.0391	0.0140	37	-1.73	130.00	0
130	L-Lysine (%)	0848	0.6000	0.0000	0.6376	0.0391	0.0140	37	-0.96	130.00	0
130	L-Lysine (%)	0682	0.6010	0.0000	0.6376	0.0391	0.0140	37	-0.94	130.00	0
130	L-Lysine (%)	0910	0.6050	0.0300	0.6376	0.0391	0.0140	37	-0.83	130.00	0
130	L-Lysine (%)	0407	0.6090	0.0300	0.6376	0.0391	0.0140	37	-0.73	130.05	0
130	L-Lysine (%)	0619	0.6195	0.0230	0.6376	0.0391	0.0140	37	-0.46	130.00	0
130	L-Lysine (%)	0684	0.6250	0.0100	0.6376	0.0391	0.0140	37	-0.32	130.00	0
130	L-Lysine (%)	0226	0.6294	0.0363	0.6376	0.0391	0.0140	37	-0.21	130.00	0
130	L-Lysine (%)	0868	0.6300	0.0040	0.6376	0.0391	0.0140	37	-0.19	130.00	0
130	L-Lysine (%)	2256	0.6315	0.0070	0.6376	0.0391	0.0140	37	-0.16	130.00	0
130	L-Lysine (%)	0872	0.6335	0.0210	0.6376	0.0391	0.0140	37	-0.11	130.00	0
130	L-Lysine (%)	0652	0.6350	0.0100	0.6376	0.0391	0.0140	37	-0.07	130.00	0
130	L-Lysine (%)	0878	0.6385	0.0170	0.6376	0.0391	0.0140	37	0.02	130.00	0
130	L-Lysine (%)	0571	0.6395	0.0070	0.6376	0.0391	0.0140	37	0.05	130.00	0
130	L-Lysine (%)	0242	0.6400	0.0400	0.6376	0.0391	0.0140	37	0.06	130.05	0
130	L-Lysine (%)	0610	0.6400	0.0000	0.6376	0.0391	0.0140	37	0.06	130.05	0
130	L-Lysine (%)	0676	0.6400	0.0000	0.6376	0.0391	0.0140	37	0.06	130.05	0
130	L-Lysine (%)	0941	0.6485	0.0050	0.6376	0.0391	0.0140	37	0.28	130.00	0
130	L-Lysine (%)	0098	0.6495	0.0050	0.6376	0.0391	0.0140	37	0.30	130.02	0
130	L-Lysine (%)	0644	0.6500	0.0100	0.6376	0.0391	0.0140	37	0.32	130.00	0
130	L-Lysine (%)	0939	0.6500	0.0000	0.6376	0.0391	0.0140	37	0.32	130.00	0
130	L-Lysine (%)	2059	0.6505	0.0030	0.6376	0.0391	0.0140	37	0.33	130.00	0
130	L-Lysine (%)	0859	0.6530	0.0125	0.6376	0.0391	0.0140	37	0.39	130.00	0
130	L-Lysine (%)	2196	0.6540	0.0000	0.6376	0.0391	0.0140	37	0.42	130.05	0
130	L-Lysine (%)	0968	0.6545	0.0230	0.6376	0.0391	0.0140	37	0.43	130.00	0
130	L-Lysine (%)	0227	0.6550	0.0100	0.6376	0.0391	0.0140	37	0.44	130.00	0
130	L-Lysine (%)	0148	0.6700	0.0200	0.6376	0.0391	0.0140	37	0.83	130.05	0
130	L-Lysine (%)	0504	0.6750	0.0100	0.6376	0.0391	0.0140	37	0.96	130.00	0
130	L-Lysine (%)	2188	0.6850	0.0100	0.6376	0.0391	0.0140	37	1.21	130.05	0
130	L-Lysine (%)	0870	0.7132	0.0161	0.6376	0.0391	0.0140	37	1.93	130.00	0
130	L-Lysine (%)	0016	0.7455	0.0170	0.6376	0.0391	0.0140	37	2.76	130.05	0
130	L-Lysine (%)	2058	0.7500	0.0000	0.6376	0.0391	0.0140	37	2.87	130.99	0
130	L-Lysine (%)	2144	0.7700	0.0200	0.6376	0.0391	0.0140	37	3.38	130.99	0
131	Methionine (%)	2246	0.1703	0.0125	0.2312	0.0206	0.0085	35	-2.96	131.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
131	Methionine (%)	0407	0.1923	0.0074	0.2312	0.0206	0.0085	35	-1.89	131.05	0
131	Methionine (%)	0353	0.1950	0.0100	0.2312	0.0206	0.0085	35	-1.76	131.00	0
131	Methionine (%)	0148	0.2000	0.0200	0.2312	0.0206	0.0085	35	-1.52	131.05	0
131	Methionine (%)	0626	0.2020	0.0120	0.2312	0.0206	0.0085	35	-1.42	131.05	0
131	Methionine (%)	0848	0.2050	0.0100	0.2312	0.0206	0.0085	35	-1.27	131.00	0
131	Methionine (%)	0675	0.2100	0.0200	0.2312	0.0206	0.0085	35	-1.03	131.00	0
131	Methionine (%)	0870	0.2116	0.0124	0.2312	0.0206	0.0085	35	-0.95	131.00	0
131	Methionine (%)	0941	0.2220	0.0020	0.2312	0.0206	0.0085	35	-0.45	131.00	0
131	Methionine (%)	0619	0.2235	0.0230	0.2312	0.0206	0.0085	35	-0.37	131.00	0
131	Methionine (%)	0684	0.2250	0.0100	0.2312	0.0206	0.0085	35	-0.30	131.00	0
131	Methionine (%)	0872	0.2250	0.0040	0.2312	0.0206	0.0085	35	-0.30	131.00	0
131	Methionine (%)	0098	0.2295	0.0050	0.2312	0.0206	0.0085	35	-0.08	131.02	0
131	Methionine (%)	0227	0.2300	0.0200	0.2312	0.0206	0.0085	35	-0.06	131.00	0
131	Methionine (%)	2144	0.2300	0.0200	0.2312	0.0206	0.0085	35	-0.06	131.99	0
131	Methionine (%)	0878	0.2340	0.0080	0.2312	0.0206	0.0085	35	0.13	131.00	0
131	Methionine (%)	0652	0.2350	0.0100	0.2312	0.0206	0.0085	35	0.18	131.00	0
131	Methionine (%)	0910	0.2350	0.0100	0.2312	0.0206	0.0085	35	0.18	131.00	0
131	Methionine (%)	0610	0.2350	0.0100	0.2312	0.0206	0.0085	35	0.18	131.05	0
131	Methionine (%)	2196	0.2350	0.0000	0.2312	0.0206	0.0085	35	0.18	131.05	0
131	Methionine (%)	0859	0.2362	0.0068	0.2312	0.0206	0.0085	35	0.24	131.00	0
131	Methionine (%)	0571	0.2385	0.0010	0.2312	0.0206	0.0085	35	0.35	131.00	0
131	Methionine (%)	0504	0.2400	0.0000	0.2312	0.0206	0.0085	35	0.43	131.00	0
131	Methionine (%)	0939	0.2400	0.0000	0.2312	0.0206	0.0085	35	0.43	131.00	0
131	Methionine (%)	0968	0.2400	0.0040	0.2312	0.0206	0.0085	35	0.43	131.00	0
131	Methionine (%)	0242	0.2400	0.0000	0.2312	0.0206	0.0085	35	0.43	131.05	0
131	Methionine (%)	0682	0.2420	0.0000	0.2312	0.0206	0.0085	35	0.52	131.00	0
131	Methionine (%)	0644	0.2445	0.0030	0.2312	0.0206	0.0085	35	0.64	131.00	0
131	Methionine (%)	2059	0.2450	0.0040	0.2312	0.0206	0.0085	35	0.67	131.00	0
131	Methionine (%)	2188	0.2500	0.0200	0.2312	0.0206	0.0085	35	0.91	131.05	0
131	Methionine (%)	0016	0.2540	0.0000	0.2312	0.0206	0.0085	35	1.11	131.05	0
131	Methionine (%)	0676	0.2550	0.0100	0.2312	0.0206	0.0085	35	1.15	131.05	0
131	Methionine (%)	0868	0.2565	0.0030	0.2312	0.0206	0.0085	35	1.23	131.00	0
131	Methionine (%)	2058	0.2600	0.0200	0.2312	0.0206	0.0085	35	1.40	131.99	0
131	Methionine (%)	0226	0.2787	0.0004	0.2312	0.0206	0.0085	35	2.31	131.00	0
131	Methionine (%)	0918	0.2300	0.0400	0.2312	0.0206	0.0085	35	-0.06	131.00	1
132	Phenylalanine (%)	0675	0.5600	0.0200	0.6719	0.0465	0.0170	35	-2.40	132.00	0
132	Phenylalanine (%)	0968	0.5820	0.0340	0.6719	0.0465	0.0170	35	-1.93	132.00	0
132	Phenylalanine (%)	2246	0.5934	0.0091	0.6719	0.0465	0.0170	35	-1.69	132.05	0
132	Phenylalanine (%)	0407	0.6055	0.0330	0.6719	0.0465	0.0170	35	-1.43	132.05	0
132	Phenylalanine (%)	0918	0.6082	0.0145	0.6719	0.0465	0.0170	35	-1.37	132.00	0
132	Phenylalanine (%)	0868	0.6290	0.0080	0.6719	0.0465	0.0170	35	-0.92	132.00	0
132	Phenylalanine (%)	0682	0.6300	0.0000	0.6719	0.0465	0.0170	35	-0.90	132.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
132	Phenylalanine (%)	0676	0.6350	0.0100	0.6719	0.0465	0.0170	35	-0.79	132.05	0
132	Phenylalanine (%)	2256	0.6385	0.0010	0.6719	0.0465	0.0170	35	-0.72	132.00	0
132	Phenylalanine (%)	0353	0.6400	0.0000	0.6719	0.0465	0.0170	35	-0.68	132.00	0
132	Phenylalanine (%)	2196	0.6430	0.0000	0.6719	0.0465	0.0170	35	-0.62	132.05	0
132	Phenylalanine (%)	0910	0.6600	0.0000	0.6719	0.0465	0.0170	35	-0.26	132.00	0
132	Phenylalanine (%)	0684	0.6650	0.0300	0.6719	0.0465	0.0170	35	-0.15	132.00	0
132	Phenylalanine (%)	0098	0.6720	0.0060	0.6719	0.0465	0.0170	35	0.00	132.02	0
132	Phenylalanine (%)	0571	0.6740	0.0040	0.6719	0.0465	0.0170	35	0.05	132.00	0
132	Phenylalanine (%)	0941	0.6745	0.0310	0.6719	0.0465	0.0170	35	0.06	132.00	0
132	Phenylalanine (%)	0242	0.6750	0.0300	0.6719	0.0465	0.0170	35	0.07	132.05	0
132	Phenylalanine (%)	0878	0.6780	0.0040	0.6719	0.0465	0.0170	35	0.13	132.00	0
132	Phenylalanine (%)	2058	0.6800	0.0600	0.6719	0.0465	0.0170	35	0.17	132.99	0
132	Phenylalanine (%)	0227	0.6800	0.0000	0.6719	0.0465	0.0170	35	0.17	132.00	0
132	Phenylalanine (%)	0226	0.6840	0.0217	0.6719	0.0465	0.0170	35	0.26	132.00	0
132	Phenylalanine (%)	0872	0.6840	0.0100	0.6719	0.0465	0.0170	35	0.26	132.00	0
132	Phenylalanine (%)	0148	0.6850	0.0300	0.6719	0.0465	0.0170	35	0.28	132.05	0
132	Phenylalanine (%)	0610	0.6900	0.0200	0.6719	0.0465	0.0170	35	0.39	132.05	0
132	Phenylalanine (%)	0626	0.6920	0.0120	0.6719	0.0465	0.0170	35	0.43	132.05	0
132	Phenylalanine (%)	0644	0.6925	0.0210	0.6719	0.0465	0.0170	35	0.44	132.00	0
132	Phenylalanine (%)	2059	0.6945	0.0050	0.6719	0.0465	0.0170	35	0.49	132.00	0
132	Phenylalanine (%)	0859	0.7065	0.0116	0.6719	0.0465	0.0170	35	0.74	132.00	0
132	Phenylalanine (%)	2188	0.7100	0.0200	0.6719	0.0465	0.0170	35	0.82	132.05	0
132	Phenylalanine (%)	0504	0.7150	0.0100	0.6719	0.0465	0.0170	35	0.93	132.00	0
132	Phenylalanine (%)	0939	0.7250	0.0100	0.6719	0.0465	0.0170	35	1.14	132.00	0
132	Phenylalanine (%)	2144	0.7300	0.0400	0.6719	0.0465	0.0170	35	1.25	132.99	0
132	Phenylalanine (%)	0652	0.7350	0.0300	0.6719	0.0465	0.0170	35	1.36	132.00	0
132	Phenylalanine (%)	0619	0.7365	0.0490	0.6719	0.0465	0.0170	35	1.39	132.00	0
132	Phenylalanine (%)	0870	0.7427	0.0102	0.6719	0.0465	0.0170	35	1.52	132.00	0
133	Proline (%)	0868	0.7790	0.0100	0.9593	0.0519	0.0220	32	-3.47	133.00	0
133	Proline (%)	2144	0.8350	0.0300	0.9593	0.0519	0.0220	32	-2.39	133.99	0
133	Proline (%)	0682	0.8590	0.0000	0.9593	0.0519	0.0220	32	-1.93	133.00	0
133	Proline (%)	2246	0.8656	0.0023	0.9593	0.0519	0.0220	32	-1.81	133.05	0
133	Proline (%)	0910	0.8950	0.0300	0.9593	0.0519	0.0220	32	-1.24	133.00	0
133	Proline (%)	2256	0.8980	0.0160	0.9593	0.0519	0.0220	32	-1.18	133.00	0
133	Proline (%)	0353	0.9050	0.0100	0.9593	0.0519	0.0220	32	-1.05	133.00	0
133	Proline (%)	0504	0.9500	0.0600	0.9593	0.0519	0.0220	32	-0.18	133.00	0
133	Proline (%)	0652	0.9500	0.0200	0.9593	0.0519	0.0220	32	-0.18	133.00	0
133	Proline (%)	0242	0.9500	0.0800	0.9593	0.0519	0.0220	32	-0.18	133.05	0
133	Proline (%)	0571	0.9520	0.0100	0.9593	0.0519	0.0220	32	-0.14	133.00	0
133	Proline (%)	0610	0.9550	0.0100	0.9593	0.0519	0.0220	32	-0.08	133.05	0
133	Proline (%)	2058	0.9550	0.0500	0.9593	0.0519	0.0220	32	-0.08	133.99	0
133	Proline (%)	0859	0.9575	0.0248	0.9593	0.0519	0.0220	32	-0.03	133.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
133	Proline (%)	0675	0.9600	0.0200	0.9593	0.0519	0.0220	32	0.01	133.00	0
133	Proline (%)	0684	0.9600	0.0000	0.9593	0.0519	0.0220	32	0.01	133.00	0
133	Proline (%)	2059	0.9605	0.0010	0.9593	0.0519	0.0220	32	0.02	133.00	0
133	Proline (%)	0407	0.9610	0.0100	0.9593	0.0519	0.0220	32	0.03	133.05	0
133	Proline (%)	0227	0.9700	0.0200	0.9593	0.0519	0.0220	32	0.21	133.00	0
133	Proline (%)	0870	0.9705	0.0023	0.9593	0.0519	0.0220	32	0.21	133.00	0
133	Proline (%)	0644	0.9710	0.0300	0.9593	0.0519	0.0220	32	0.23	133.00	0
133	Proline (%)	0878	0.9725	0.0090	0.9593	0.0519	0.0220	32	0.25	133.00	0
133	Proline (%)	0941	0.9800	0.0300	0.9593	0.0519	0.0220	32	0.40	133.00	0
133	Proline (%)	0872	0.9815	0.0030	0.9593	0.0519	0.0220	32	0.43	133.00	0
133	Proline (%)	0939	0.9850	0.0100	0.9593	0.0519	0.0220	32	0.49	133.00	0
133	Proline (%)	0626	0.9920	0.0100	0.9593	0.0519	0.0220	32	0.63	133.05	0
133	Proline (%)	0148	0.9950	0.0100	0.9593	0.0519	0.0220	32	0.69	133.05	0
133	Proline (%)	2196	0.9970	0.0000	0.9593	0.0519	0.0220	32	0.73	133.05	0
133	Proline (%)	0968	1.055	0.0130	0.9593	0.0519	0.0220	32	1.83	133.00	0
133	Proline (%)	2188	1.090	0.0400	0.9593	0.0519	0.0220	32	2.52	133.05	0
133	Proline (%)	0619	1.170	0.0000	0.9593	0.0519	0.0220	32	4.06	133.00	0
133	Proline (%)	0226	1.504	0.1426	0.9593	0.0519	0.0220	32	10.49	133.00	0
133	Proline (%)	0918	0.9000	0.1940	0.9593	0.0519	0.0220	32	-1.14	133.00	1
134	Serine (%)	0504	0.5700	0.0200	0.6898	0.0388	0.0161	35	-3.09	134.00	0
134	Serine (%)	0098	0.6290	0.0240	0.6898	0.0388	0.0161	35	-1.57	134.02	0
134	Serine (%)	0610	0.6450	0.0100	0.6898	0.0388	0.0161	35	-1.15	134.05	0
134	Serine (%)	0682	0.6490	0.0000	0.6898	0.0388	0.0161	35	-1.05	134.00	0
134	Serine (%)	0675	0.6500	0.0200	0.6898	0.0388	0.0161	35	-1.03	134.00	0
134	Serine (%)	0918	0.6574	0.0354	0.6898	0.0388	0.0161	35	-0.83	134.00	0
134	Serine (%)	0676	0.6600	0.0400	0.6898	0.0388	0.0161	35	-0.77	134.05	0
134	Serine (%)	0644	0.6630	0.0020	0.6898	0.0388	0.0161	35	-0.69	134.00	0
134	Serine (%)	0242	0.6650	0.0500	0.6898	0.0388	0.0161	35	-0.64	134.05	0
134	Serine (%)	0226	0.6655	0.0268	0.6898	0.0388	0.0161	35	-0.63	134.00	0
134	Serine (%)	2059	0.6660	0.0100	0.6898	0.0388	0.0161	35	-0.61	134.00	0
134	Serine (%)	2256	0.6735	0.0050	0.6898	0.0388	0.0161	35	-0.42	134.00	0
134	Serine (%)	0652	0.6750	0.0100	0.6898	0.0388	0.0161	35	-0.38	134.00	0
134	Serine (%)	0407	0.6755	0.0350	0.6898	0.0388	0.0161	35	-0.37	134.05	0
134	Serine (%)	0626	0.6755	0.0050	0.6898	0.0388	0.0161	35	-0.37	134.05	0
134	Serine (%)	0910	0.6800	0.0200	0.6898	0.0388	0.0161	35	-0.25	134.00	0
134	Serine (%)	0571	0.6845	0.0030	0.6898	0.0388	0.0161	35	-0.14	134.00	0
134	Serine (%)	2144	0.6850	0.0500	0.6898	0.0388	0.0161	35	-0.12	134.99	0
134	Serine (%)	0227	0.6900	0.0000	0.6898	0.0388	0.0161	35	0.01	134.00	0
134	Serine (%)	2246	0.6938	0.0043	0.6898	0.0388	0.0161	35	0.10	134.05	0
134	Serine (%)	0353	0.6950	0.0100	0.6898	0.0388	0.0161	35	0.14	134.00	0
134	Serine (%)	0968	0.6955	0.0070	0.6898	0.0388	0.0161	35	0.15	134.00	0
134	Serine (%)	0872	0.6985	0.0110	0.6898	0.0388	0.0161	35	0.23	134.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
134	Serine (%)	0859	0.6992	0.0135	0.6898	0.0388	0.0161	35	0.24	134.00	0
134	Serine (%)	0941	0.7035	0.0150	0.6898	0.0388	0.0161	35	0.35	134.00	0
134	Serine (%)	0878	0.7060	0.0100	0.6898	0.0388	0.0161	35	0.42	134.00	0
134	Serine (%)	0868	0.7075	0.0030	0.6898	0.0388	0.0161	35	0.46	134.00	0
134	Serine (%)	0684	0.7200	0.0000	0.6898	0.0388	0.0161	35	0.78	134.00	0
134	Serine (%)	0148	0.7200	0.0400	0.6898	0.0388	0.0161	35	0.78	134.05	0
134	Serine (%)	0939	0.7400	0.0000	0.6898	0.0388	0.0161	35	1.30	134.00	0
134	Serine (%)	2188	0.7550	0.0300	0.6898	0.0388	0.0161	35	1.68	134.05	0
134	Serine (%)	0619	0.7715	0.0010	0.6898	0.0388	0.0161	35	2.11	134.00	0
134	Serine (%)	2196	0.7720	0.0000	0.6898	0.0388	0.0161	35	2.12	134.05	0
134	Serine (%)	0870	0.7778	0.0315	0.6898	0.0388	0.0161	35	2.27	134.00	0
134	Serine (%)	2058	0.9300	0.0200	0.6898	0.0388	0.0161	35	6.20	134.99	0
135	Threonine (%)	0407	0.4255	0.0310	0.5136	0.0225	0.0136	35	-3.92	135.05	0
135	Threonine (%)	0626	0.4580	0.0160	0.5136	0.0225	0.0136	35	-2.47	135.05	0
135	Threonine (%)	0675	0.4600	0.0200	0.5136	0.0225	0.0136	35	-2.38	135.00	0
135	Threonine (%)	0682	0.4690	0.0000	0.5136	0.0225	0.0136	35	-1.98	135.00	0
135	Threonine (%)	0226	0.4902	0.0294	0.5136	0.0225	0.0136	35	-1.04	135.00	0
135	Threonine (%)	0353	0.4950	0.0100	0.5136	0.0225	0.0136	35	-0.83	135.00	0
135	Threonine (%)	0676	0.4950	0.0300	0.5136	0.0225	0.0136	35	-0.83	135.05	0
135	Threonine (%)	2144	0.4950	0.0300	0.5136	0.0225	0.0136	35	-0.83	135.99	0
135	Threonine (%)	0652	0.5000	0.0000	0.5136	0.0225	0.0136	35	-0.60	135.00	0
135	Threonine (%)	2188	0.5000	0.0200	0.5136	0.0225	0.0136	35	-0.60	135.05	0
135	Threonine (%)	0504	0.5050	0.0100	0.5136	0.0225	0.0136	35	-0.38	135.00	0
135	Threonine (%)	0910	0.5050	0.0100	0.5136	0.0225	0.0136	35	-0.38	135.00	0
135	Threonine (%)	0610	0.5100	0.0000	0.5136	0.0225	0.0136	35	-0.16	135.05	0
135	Threonine (%)	0872	0.5115	0.0030	0.5136	0.0225	0.0136	35	-0.09	135.00	0
135	Threonine (%)	2256	0.5115	0.0050	0.5136	0.0225	0.0136	35	-0.09	135.00	0
135	Threonine (%)	0918	0.5127	0.0326	0.5136	0.0225	0.0136	35	-0.04	135.00	0
135	Threonine (%)	2059	0.5130	0.0060	0.5136	0.0225	0.0136	35	-0.03	135.00	0
135	Threonine (%)	0619	0.5135	0.0210	0.5136	0.0225	0.0136	35	0.00	135.00	0
135	Threonine (%)	0571	0.5145	0.0090	0.5136	0.0225	0.0136	35	0.04	135.00	0
135	Threonine (%)	2246	0.5153	0.0149	0.5136	0.0225	0.0136	35	0.07	135.05	0
135	Threonine (%)	0644	0.5170	0.0020	0.5136	0.0225	0.0136	35	0.15	135.00	0
135	Threonine (%)	0878	0.5185	0.0130	0.5136	0.0225	0.0136	35	0.22	135.00	0
135	Threonine (%)	0968	0.5190	0.0040	0.5136	0.0225	0.0136	35	0.24	135.00	0
135	Threonine (%)	0868	0.5205	0.0010	0.5136	0.0225	0.0136	35	0.31	135.00	0
135	Threonine (%)	0859	0.5237	0.0127	0.5136	0.0225	0.0136	35	0.45	135.00	0
135	Threonine (%)	0227	0.5250	0.0100	0.5136	0.0225	0.0136	35	0.51	135.00	0
135	Threonine (%)	0941	0.5250	0.0120	0.5136	0.0225	0.0136	35	0.51	135.00	0
135	Threonine (%)	0148	0.5250	0.0100	0.5136	0.0225	0.0136	35	0.51	135.05	0
135	Threonine (%)	0098	0.5265	0.0110	0.5136	0.0225	0.0136	35	0.57	135.02	0
135	Threonine (%)	0242	0.5350	0.0300	0.5136	0.0225	0.0136	35	0.95	135.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
135	Threonine (%)	0939	0.5450	0.0100	0.5136	0.0225	0.0136	35	1.40	135.00	0
135	Threonine (%)	2196	0.5510	0.0000	0.5136	0.0225	0.0136	35	1.66	135.05	0
135	Threonine (%)	0684	0.5550	0.0100	0.5136	0.0225	0.0136	35	1.84	135.00	0
135	Threonine (%)	0870	0.5565	0.0207	0.5136	0.0225	0.0136	35	1.90	135.00	0
135	Threonine (%)	2058	0.7050	0.0300	0.5136	0.0225	0.0136	35	8.51	135.99	0
136	Tryptophan (%)	0407	0.1130	0.0080	0.2037	0.0258	0.0046	23	-3.52	136.05	0
136	Tryptophan (%)	2246	0.1136	0.0070	0.2037	0.0258	0.0046	23	-3.49	136.05	0
136	Tryptophan (%)	0504	0.1250	0.0100	0.2037	0.0258	0.0046	23	-3.05	136.99	0
136	Tryptophan (%)	0870	0.1683	0.0084	0.2037	0.0258	0.0046	23	-1.37	136.00	0
136	Tryptophan (%)	0868	0.1875	0.0010	0.2037	0.0258	0.0046	23	-0.63	136.01	0
136	Tryptophan (%)	2058	0.1900	0.0000	0.2037	0.0258	0.0046	23	-0.53	136.03	0
136	Tryptophan (%)	0878	0.1960	0.0040	0.2037	0.0258	0.0046	23	-0.30	136.01	0
136	Tryptophan (%)	0242	0.2000	0.0000	0.2037	0.0258	0.0046	23	-0.14	136.05	0
136	Tryptophan (%)	0939	0.2000	0.0000	0.2037	0.0258	0.0046	23	-0.14	136.99	0
136	Tryptophan (%)	0872	0.2050	0.0020	0.2037	0.0258	0.0046	23	0.05	136.03	0
136	Tryptophan (%)	0918	0.2050	0.0100	0.2037	0.0258	0.0046	23	0.05	136.00	0
136	Tryptophan (%)	0859	0.2060	0.0081	0.2037	0.0258	0.0046	23	0.09	136.03	0
136	Tryptophan (%)	0098	0.2080	0.0020	0.2037	0.0258	0.0046	23	0.17	136.02	0
136	Tryptophan (%)	0148	0.2100	0.0000	0.2037	0.0258	0.0046	23	0.24	136.03	0
136	Tryptophan (%)	2059	0.2100	0.0000	0.2037	0.0258	0.0046	23	0.24	136.03	0
136	Tryptophan (%)	0619	0.2110	0.0020	0.2037	0.0258	0.0046	23	0.28	136.00	0
136	Tryptophan (%)	0941	0.2115	0.0010	0.2037	0.0258	0.0046	23	0.30	136.00	0
136	Tryptophan (%)	0644	0.2170	0.0020	0.2037	0.0258	0.0046	23	0.51	136.01	0
136	Tryptophan (%)	0571	0.2205	0.0010	0.2037	0.0258	0.0046	23	0.65	136.03	0
136	Tryptophan (%)	0682	0.2250	0.0000	0.2037	0.0258	0.0046	23	0.82	136.00	0
136	Tryptophan (%)	0684	0.2350	0.0100	0.2037	0.0258	0.0046	23	1.21	136.00	0
136	Tryptophan (%)	0353	0.2450	0.0100	0.2037	0.0258	0.0046	23	1.60	136.03	0
136	Tryptophan (%)	0227	0.2500	0.0200	0.2037	0.0258	0.0046	23	1.79	136.00	0
136	Tryptophan (%)	0910	0.1950	0.0300	0.2037	0.0258	0.0046	23	-0.34	136.00	1
137	Tyrosine (%)	0675	0.3200	0.0200	0.4417	0.0652	0.0155	28	-1.87	137.00	0
137	Tyrosine (%)	0682	0.3390	0.0000	0.4417	0.0652	0.0155	28	-1.57	137.00	0
137	Tyrosine (%)	2144	0.3500	0.0000	0.4417	0.0652	0.0155	28	-1.41	137.99	0
137	Tyrosine (%)	0968	0.3690	0.0120	0.4417	0.0652	0.0155	28	-1.11	137.00	0
137	Tyrosine (%)	0910	0.3700	0.0000	0.4417	0.0652	0.0155	28	-1.10	137.00	0
137	Tyrosine (%)	0918	0.3726	0.0077	0.4417	0.0652	0.0155	28	-1.06	137.00	0
137	Tyrosine (%)	2246	0.3726	0.0115	0.4417	0.0652	0.0155	28	-1.06	137.05	0
137	Tyrosine (%)	0227	0.4150	0.0300	0.4417	0.0652	0.0155	28	-0.41	137.00	0
137	Tyrosine (%)	0098	0.4210	0.0060	0.4417	0.0652	0.0155	28	-0.32	137.02	0
137	Tyrosine (%)	0941	0.4230	0.0300	0.4417	0.0652	0.0155	28	-0.29	137.00	0
137	Tyrosine (%)	0504	0.4350	0.0100	0.4417	0.0652	0.0155	28	-0.10	137.00	0
137	Tyrosine (%)	0242	0.4400	0.0400	0.4417	0.0652	0.0155	28	-0.03	137.05	0
137	Tyrosine (%)	2256	0.4445	0.0010	0.4417	0.0652	0.0155	28	0.04	137.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
137	Tyrosine (%)	0872	0.4455	0.0310	0.4417	0.0652	0.0155	28	0.06	137.00	0
137	Tyrosine (%)	0644	0.4525	0.0150	0.4417	0.0652	0.0155	28	0.17	137.00	0
137	Tyrosine (%)	0226	0.4623	0.0070	0.4417	0.0652	0.0155	28	0.32	137.00	0
137	Tyrosine (%)	0353	0.4650	0.0100	0.4417	0.0652	0.0155	28	0.36	137.00	0
137	Tyrosine (%)	0626	0.4660	0.0320	0.4417	0.0652	0.0155	28	0.37	137.05	0
137	Tyrosine (%)	0684	0.4700	0.0200	0.4417	0.0652	0.0155	28	0.43	137.00	0
137	Tyrosine (%)	2059	0.4735	0.0070	0.4417	0.0652	0.0155	28	0.49	137.00	0
137	Tyrosine (%)	0870	0.4762	0.0148	0.4417	0.0652	0.0155	28	0.53	137.00	0
137	Tyrosine (%)	2196	0.4770	0.0000	0.4417	0.0652	0.0155	28	0.54	137.05	0
137	Tyrosine (%)	0939	0.4800	0.0200	0.4417	0.0652	0.0155	28	0.59	137.00	0
137	Tyrosine (%)	0676	0.4850	0.0300	0.4417	0.0652	0.0155	28	0.66	137.05	0
137	Tyrosine (%)	0610	0.5050	0.0100	0.4417	0.0652	0.0155	28	0.97	137.05	0
137	Tyrosine (%)	0148	0.5350	0.0300	0.4417	0.0652	0.0155	28	1.43	137.05	0
137	Tyrosine (%)	2188	0.5350	0.0100	0.4417	0.0652	0.0155	28	1.43	137.05	0
137	Tyrosine (%)	2058	0.5850	0.0300	0.4417	0.0652	0.0155	28	2.20	137.99	0
137	Tyrosine (%)	0407	0.3960	0.0880	0.4417	0.0652	0.0155	28	-0.70	137.05	1
138	Valine (%)	0226	0.4966	0.0653	0.6987	0.0499	0.0166	34	-4.05	138.00	0
138	Valine (%)	0353	0.5650	0.0100	0.6987	0.0499	0.0166	34	-2.68	138.00	0
138	Valine (%)	0619	0.5785	0.0510	0.6987	0.0499	0.0166	34	-2.41	138.00	0
138	Valine (%)	2256	0.6255	0.0050	0.6987	0.0499	0.0166	34	-1.47	138.00	0
138	Valine (%)	0407	0.6320	0.0340	0.6987	0.0499	0.0166	34	-1.34	138.05	0
138	Valine (%)	0675	0.6400	0.0200	0.6987	0.0499	0.0166	34	-1.18	138.00	0
138	Valine (%)	0918	0.6507	0.0166	0.6987	0.0499	0.0166	34	-0.96	138.00	0
138	Valine (%)	0868	0.6640	0.0020	0.6987	0.0499	0.0166	34	-0.70	138.00	0
138	Valine (%)	2196	0.6670	0.0000	0.6987	0.0499	0.0166	34	-0.64	138.05	0
138	Valine (%)	0682	0.6720	0.0000	0.6987	0.0499	0.0166	34	-0.54	138.00	0
138	Valine (%)	0859	0.6792	0.0084	0.6987	0.0499	0.0166	34	-0.39	138.00	0
138	Valine (%)	0684	0.6800	0.0600	0.6987	0.0499	0.0166	34	-0.38	138.00	0
138	Valine (%)	0968	0.6865	0.0170	0.6987	0.0499	0.0166	34	-0.25	138.00	0
138	Valine (%)	0676	0.6950	0.0100	0.6987	0.0499	0.0166	34	-0.08	138.05	0
138	Valine (%)	0910	0.6950	0.0300	0.6987	0.0499	0.0166	34	-0.08	138.00	0
138	Valine (%)	0242	0.7000	0.0200	0.6987	0.0499	0.0166	34	0.03	138.05	0
138	Valine (%)	0872	0.7065	0.0270	0.6987	0.0499	0.0166	34	0.16	138.00	0
138	Valine (%)	2246	0.7077	0.0080	0.6987	0.0499	0.0166	34	0.18	138.05	0
138	Valine (%)	0504	0.7150	0.0100	0.6987	0.0499	0.0166	34	0.33	138.00	0
138	Valine (%)	0878	0.7150	0.0020	0.6987	0.0499	0.0166	34	0.33	138.00	0
138	Valine (%)	0939	0.7150	0.0100	0.6987	0.0499	0.0166	34	0.33	138.00	0
138	Valine (%)	0571	0.7190	0.0060	0.6987	0.0499	0.0166	34	0.41	138.00	0
138	Valine (%)	0652	0.7200	0.0200	0.6987	0.0499	0.0166	34	0.43	138.00	0
138	Valine (%)	0098	0.7290	0.0080	0.6987	0.0499	0.0166	34	0.61	138.02	0
138	Valine (%)	2059	0.7300	0.0040	0.6987	0.0499	0.0166	34	0.63	138.00	0
138	Valine (%)	0626	0.7355	0.0110	0.6987	0.0499	0.0166	34	0.74	138.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
138	Valine (%)	0227	0.7400	0.0200	0.6987	0.0499	0.0166	34	0.83	138.00	0
138	Valine (%)	2188	0.7400	0.0000	0.6987	0.0499	0.0166	34	0.83	138.05	0
138	Valine (%)	0870	0.7405	0.0337	0.6987	0.0499	0.0166	34	0.84	138.00	0
138	Valine (%)	0148	0.7450	0.0100	0.6987	0.0499	0.0166	34	0.93	138.05	0
138	Valine (%)	0644	0.7535	0.0050	0.6987	0.0499	0.0166	34	1.10	138.00	0
138	Valine (%)	0941	0.7535	0.0010	0.6987	0.0499	0.0166	34	1.10	138.00	0
138	Valine (%)	0610	0.7600	0.0200	0.6987	0.0499	0.0166	34	1.23	138.05	0
138	Valine (%)	2058	0.8200	0.0200	0.6987	0.0499	0.0166	34	2.43	138.99	0
139	Taurine (%)	0407	0.0095	0.0010				3		139.05	0
139	Taurine (%)	0682	0.0600	0.0000				3		139.00	0
139	Taurine (%)	0504	0.1200	0.0000				3		139.00	0
139	Taurine (%)	0941	0.0930	0.0040				3		139.00	1
139	Taurine (%)	0098	< 0.01					3		139.02	5
160	Fructose (%)	0297	0.1245	0.0010				2		160.99	0
160	Fructose (%)	0096	0.1590	0.0169				2		160.99	0
160	Fructose (%)	0227	< 0.15					2		160.99	5
161	Galactose (%)	0297	0.0000	0.0000				0		161.99	4
162	Glucose (%)	0297	0.0860	0.0020				2		162.99	0
162	Glucose (%)	0227	0.2600	0.0400				2		162.99	0
162	Glucose (%)	0096	< 0.2					2		162.99	5
163	Lactose (%)	0297	0.0000	0.0000				0		163.99	4
163	Lactose (%)	0227	< 0.15					0		163.99	5
163	Lactose (%)	0096	< 0.4					0		163.99	5
164	Maltose (%)	0297	0.2855	0.0530				3		164.99	0
164	Maltose (%)	0227	0.3450	0.0300				3		164.99	0
164	Maltose (%)	0096	0.4803	0.0233				3		164.99	0
165	Sucrose (%)	0096	1.212	0.2688				3		165.99	0
165	Sucrose (%)	0227	1.245	0.0300				3		165.99	0
165	Sucrose (%)	0297	1.743	0.0060				3		165.99	0
166	Raffinose (%)	0297	0.2835	0.0290				2		166.99	0
166	Raffinose (%)	0227	0.6700	0.0600				2		166.99	0
167	Stachyose (%)	0227	0.1105	0.1210				2		167.99	0
167	Stachyose (%)	0297	0.4195	0.0350				2		167.99	0
348	Bacitracin (mg/kg (ppm))	2246	0.0414	0.0000				1		348.99	0
351	Chlortetracycline (mg/kg (ppm))	2248	1.415	0.1700	2.540	0.8929	0.2281	4		351.05	0
351	Chlortetracycline (mg/kg (ppm))	0939	2.424	0.0500	2.540	0.8929	0.2281	4		351.05	0
351	Chlortetracycline (mg/kg (ppm))	0218	2.744	0.1225	2.540	0.8929	0.2281	4		351.00	0
351	Chlortetracycline (mg/kg (ppm))	2135	3.575	0.5700	2.540	0.8929	0.2281	4		351.05	0
351	Chlortetracycline (mg/kg (ppm))	0910	4.000	2.000	2.540	0.8929	0.2281	4		351.03	1
354	Decoquinatate (mg/kg (ppm))	2248	2.495	0.1700	42.21	4.487	0.9027	13	-8.85	354.04	0
354	Decoquinatate (mg/kg (ppm))	2053	36.60	1.000	42.21	4.487	0.9027	13	-1.25	354.03	0
354	Decoquinatate (mg/kg (ppm))	0610	37.29	3.020	42.21	4.487	0.9027	13	-1.10	354.04	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
354	Decoquinat (mg/kg (ppm))	0047	38.70	0.6000	42.21	4.487	0.9027	13	-0.78	354.01	0
354	Decoquinat (mg/kg (ppm))	0033	41.10	1.000	42.21	4.487	0.9027	13	-0.25	354.02	0
354	Decoquinat (mg/kg (ppm))	0910	42.15	1.100	42.21	4.487	0.9027	13	-0.01	354.01	0
354	Decoquinat (mg/kg (ppm))	0001	44.25	1.500	42.21	4.487	0.9027	13	0.45	354.02	0
354	Decoquinat (mg/kg (ppm))	0032	44.75	0.1000	42.21	4.487	0.9027	13	0.57	354.01	0
354	Decoquinat (mg/kg (ppm))	0013	45.05	0.3000	42.21	4.487	0.9027	13	0.63	354.02	0
354	Decoquinat (mg/kg (ppm))	0028	45.15	0.1000	42.21	4.487	0.9027	13	0.65	354.01	0
354	Decoquinat (mg/kg (ppm))	0036	45.30	0.8000	42.21	4.487	0.9027	13	0.69	354.02	0
354	Decoquinat (mg/kg (ppm))	0218	46.20	0.9457	42.21	4.487	0.9027	13	0.89	354.01	0
354	Decoquinat (mg/kg (ppm))	0010	46.75	1.100	42.21	4.487	0.9027	13	1.01	354.01	0
354	Decoquinat (mg/kg (ppm))	2113	50.00	8.000	42.21	4.487	0.9027	13	1.74	354.04	1
365	Monensin (mg/kg (ppm))	2248	11.80	2.170	23.02	2.987	0.9337	25	-3.76	365.05	0
365	Monensin (mg/kg (ppm))	0013	16.23	2.154	23.02	2.987	0.9337	25	-2.27	365.04	0
365	Monensin (mg/kg (ppm))	2055	18.50	3.000	23.02	2.987	0.9337	25	-1.51	365.05	0
365	Monensin (mg/kg (ppm))	2135	20.11	0.9020	23.02	2.987	0.9337	25	-0.97	365.02	0
365	Monensin (mg/kg (ppm))	2234	20.43	1.910	23.02	2.987	0.9337	25	-0.87	365.05	0
365	Monensin (mg/kg (ppm))	2113	20.50	1.000	23.02	2.987	0.9337	25	-0.84	365.05	0
365	Monensin (mg/kg (ppm))	0610	20.90	0.2000	23.02	2.987	0.9337	25	-0.71	365.02	0
365	Monensin (mg/kg (ppm))	0910	21.75	0.7000	23.02	2.987	0.9337	25	-0.42	365.99	0
365	Monensin (mg/kg (ppm))	0047	22.05	0.1000	23.02	2.987	0.9337	25	-0.32	365.03	0
365	Monensin (mg/kg (ppm))	0035	22.63	0.0300	23.02	2.987	0.9337	25	-0.13	365.03	0
365	Monensin (mg/kg (ppm))	0035	22.79	0.5400	23.02	2.987	0.9337	25	-0.08	365.00	0
365	Monensin (mg/kg (ppm))	0001	22.90	0.8000	23.02	2.987	0.9337	25	-0.04	365.05	0
365	Monensin (mg/kg (ppm))	0036	23.00	0.2000	23.02	2.987	0.9337	25	-0.01	365.03	0
365	Monensin (mg/kg (ppm))	0610	23.28	2.530	23.02	2.987	0.9337	25	0.09	365.05	0
365	Monensin (mg/kg (ppm))	0010	23.75	0.3000	23.02	2.987	0.9337	25	0.25	365.03	0
365	Monensin (mg/kg (ppm))	2053	24.25	0.5000	23.02	2.987	0.9337	25	0.41	365.04	0
365	Monensin (mg/kg (ppm))	0668	24.25	0.3000	23.02	2.987	0.9337	25	0.41	365.99	0
365	Monensin (mg/kg (ppm))	0218	24.44	0.7370	23.02	2.987	0.9337	25	0.47	365.03	0
365	Monensin (mg/kg (ppm))	0148	25.00	0.1100	23.02	2.987	0.9337	25	0.66	365.03	0
365	Monensin (mg/kg (ppm))	0668	25.55	1.300	23.02	2.987	0.9337	25	0.85	365.02	0
365	Monensin (mg/kg (ppm))	0512	25.57	0.8500	23.02	2.987	0.9337	25	0.85	365.99	0
365	Monensin (mg/kg (ppm))	0033	26.00	0.4000	23.02	2.987	0.9337	25	1.00	365.03	0
365	Monensin (mg/kg (ppm))	0032	26.20	0.6000	23.02	2.987	0.9337	25	1.07	365.03	0
365	Monensin (mg/kg (ppm))	0553	27.02	1.910	23.02	2.987	0.9337	25	1.34	365.05	0
365	Monensin (mg/kg (ppm))	0171	29.05	0.1000	23.02	2.987	0.9337	25	2.02	365.99	0
365	Monensin (mg/kg (ppm))	0028	64.70	2.400	23.02	2.987	0.9337	25	13.96	365.05	2
382	Sulfamethazine (mg/kg (ppm))	2234	2.275	0.0700				3		382.04	0
382	Sulfamethazine (mg/kg (ppm))	2248	2.475	0.2300				3		382.04	0
382	Sulfamethazine (mg/kg (ppm))	0218	2.948	0.0749				3		382.02	0
388	Tylosin (mg/kg (ppm))	2234	0.1425	0.0210				3		388.05	0
388	Tylosin (mg/kg (ppm))	2135	0.3950	0.0700				3		388.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
388	Tylosin (mg/kg (ppm))	2248	2.315	0.0100				3		388.05	0
388	Tylosin (mg/kg (ppm))	0910	< 0.6					3		388.03	5
389	Virginiamycin (mg/kg (ppm))	2246	40.23	1.300				1		389.99	0
400	Water Activity (Units)	0407	0.4331	0.0017	0.4543	0.0195	0.0072	7	-1.09	400.01	0
400	Water Activity (Units)	0942	0.4360	0.0080	0.4543	0.0195	0.0072	7	-0.94	400.01	0
400	Water Activity (Units)	0589	0.4495	0.0150	0.4543	0.0195	0.0072	7	-0.25	400.01	0
400	Water Activity (Units)	2181	0.4555	0.0055	0.4543	0.0195	0.0072	7	0.06	400.01	0
400	Water Activity (Units)	2073	0.4585	0.0050	0.4543	0.0195	0.0072	7	0.21	400.99	0
400	Water Activity (Units)	2033	0.4640	0.0060	0.4543	0.0195	0.0072	7	0.50	400.99	0
400	Water Activity (Units)	2109	0.5404	0.0090	0.4543	0.0195	0.0072	7	4.41	400.01	0
412	Starch, Dietary (%)	0353	15.45	0.1500				1		412.01	0
516	Arsenic, Total (mg / kg (ppm))	0045	0.1520	0.0360	0.2095	0.0737	0.0254	8	-0.78	516.00	0
516	Arsenic, Total (mg / kg (ppm))	0186	0.1550	0.0000	0.2095	0.0737	0.0254	8	-0.74	516.52	0
516	Arsenic, Total (mg / kg (ppm))	0910	0.1600	0.0400	0.2095	0.0737	0.0254	8	-0.67	516.52	0
516	Arsenic, Total (mg / kg (ppm))	0227	0.1870	0.0160	0.2095	0.0737	0.0254	8	-0.30	516.53	0
516	Arsenic, Total (mg / kg (ppm))	0171	0.1965	0.0650	0.2095	0.0737	0.0254	8	-0.18	516.00	0
516	Arsenic, Total (mg / kg (ppm))	2053	0.2065	0.0050	0.2095	0.0737	0.0254	8	-0.04	516.52	0
516	Arsenic, Total (mg / kg (ppm))	0918	0.2987	0.0016	0.2095	0.0737	0.0254	8	1.21	516.53	0
516	Arsenic, Total (mg / kg (ppm))	0407	1.578	0.0397	0.2095	0.0737	0.0254	8	18.56	516.43	0
518	Cadmium (mg / kg (ppm))	0619	0.0650	0.0020	0.1266	0.0324	0.0084	11	-1.90	518.41	0
518	Cadmium (mg / kg (ppm))	2246	0.0790	0.0060	0.1266	0.0324	0.0084	11	-1.47	518.34	0
518	Cadmium (mg / kg (ppm))	0695	0.1060	0.0100	0.1266	0.0324	0.0084	11	-0.64	518.41	0
518	Cadmium (mg / kg (ppm))	0910	0.1200	0.0200	0.1266	0.0324	0.0084	11	-0.20	518.52	0
518	Cadmium (mg / kg (ppm))	0407	0.1214	0.0016	0.1266	0.0324	0.0084	11	-0.16	518.41	0
518	Cadmium (mg / kg (ppm))	0227	0.1395	0.0230	0.1266	0.0324	0.0084	11	0.40	518.53	0
518	Cadmium (mg / kg (ppm))	0186	0.1410	0.0140	0.1266	0.0324	0.0084	11	0.45	518.52	0
518	Cadmium (mg / kg (ppm))	0968	0.1452	0.0086	0.1266	0.0324	0.0084	11	0.57	518.43	0
518	Cadmium (mg / kg (ppm))	0918	0.1509	0.0022	0.1266	0.0324	0.0084	11	0.75	518.53	0
518	Cadmium (mg / kg (ppm))	0171	0.1530	0.0000	0.1266	0.0324	0.0084	11	0.82	518.41	0
518	Cadmium (mg / kg (ppm))	2053	0.1585	0.0050	0.1266	0.0324	0.0084	11	0.99	518.52	0
518	Cadmium (mg / kg (ppm))	0689	< 0.2		0.1266	0.0324	0.0084	11		518.31	5
520	Chromium (mg / kg (ppm))	0186	2.510	0.4000	6.596	3.840	0.4424	10	-1.06	520.52	0
520	Chromium (mg / kg (ppm))	0619	3.435	0.4500	6.596	3.840	0.4424	10	-0.82	520.41	0
520	Chromium (mg / kg (ppm))	0968	3.934	0.1117	6.596	3.840	0.4424	10	-0.69	520.43	0
520	Chromium (mg / kg (ppm))	0918	3.965	0.6737	6.596	3.840	0.4424	10	-0.69	520.53	0
520	Chromium (mg / kg (ppm))	0407	5.079	0.3575	6.596	3.840	0.4424	10	-0.40	520.41	0
520	Chromium (mg / kg (ppm))	0171	5.471	1.207	6.596	3.840	0.4424	10	-0.29	520.41	0
520	Chromium (mg / kg (ppm))	0693	9.539	0.3140	6.596	3.840	0.4424	10	0.77	520.42	0
520	Chromium (mg / kg (ppm))	0045	10.40	0.8000	6.596	3.840	0.4424	10	0.99	520.42	0
520	Chromium (mg / kg (ppm))	0164	10.45	0.1000	6.596	3.840	0.4424	10	1.00	520.53	0
520	Chromium (mg / kg (ppm))	0510	11.18	0.0100	6.596	3.840	0.4424	10	1.19	520.43	0
526	Lead (mg / kg (ppm))	0619	0.0285	0.0030	0.1425	0.0449	0.0299	9	-2.54	526.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
526	Lead (mg / kg (ppm))	0910	0.1050	0.0300	0.1425	0.0449	0.0299	9	-0.84	526.52	0
526	Lead (mg / kg (ppm))	0918	0.1331	0.0005	0.1425	0.0449	0.0299	9	-0.21	526.53	0
526	Lead (mg / kg (ppm))	0227	0.1370	0.0100	0.1425	0.0449	0.0299	9	-0.12	526.53	0
526	Lead (mg / kg (ppm))	0186	0.1380	0.0180	0.1425	0.0449	0.0299	9	-0.10	526.52	0
526	Lead (mg / kg (ppm))	0171	0.1450	0.0420	0.1425	0.0449	0.0299	9	0.06	526.41	0
526	Lead (mg / kg (ppm))	0407	0.1531	0.0060	0.1425	0.0449	0.0299	9	0.24	526.41	0
526	Lead (mg / kg (ppm))	2053	0.1865	0.0510	0.1425	0.0449	0.0299	9	0.98	526.52	0
526	Lead (mg / kg (ppm))	2246	0.5772	0.1089	0.1425	0.0449	0.0299	9	9.68	526.34	0
526	Lead (mg / kg (ppm))	0689	< 0.2		0.1425	0.0449	0.0299	9		526.31	5
539	Nickel (mg / kg (ppm))	0186	1.805	0.2100	3.234	1.349	0.1696	4		539.52	0
539	Nickel (mg / kg (ppm))	0918	2.438	0.2377	3.234	1.349	0.1696	4		539.53	0
539	Nickel (mg / kg (ppm))	0407	3.951	0.2236	3.234	1.349	0.1696	4		539.41	0
539	Nickel (mg / kg (ppm))	0171	4.741	0.0070	3.234	1.349	0.1696	4		539.41	0
710	Lauric Acid (12:0) (% (w/w))	0226	0.0013	0.0000				1		710.99	0
710	Lauric Acid (12:0) (% (w/w))	0164	< 0.02					1		710.99	5
714	Myristic Acid (14:0) (% (w/w))	0226	0.0063	0.0004				1		714.99	0
714	Myristic Acid (14:0) (% (w/w))	0164	< 0.02					1		714.99	5
716	Palmitic Acid (16:0) (% (w/w))	0226	0.6133	0.0122				2		716.99	0
716	Palmitic Acid (16:0) (% (w/w))	0164	0.6650	0.0700				2		716.99	0
718	Palmitoleic Acid (9c-16:1) (% (w/w))	0226	0.0075	0.0001				1		718.99	0
718	Palmitoleic Acid (9c-16:1) (% (w/w))	0164	< 0.04					1		718.99	5
722	Stearic Acid (18:0) (% (w/w))	0226	0.0733	0.0012				2		722.99	0
722	Stearic Acid (18:0) (% (w/w))	0164	0.0850	0.0100				2		722.99	0
724	Oleic Acid (9c-18:1) (% (w/w))	0226	0.6983	0.0108				2		724.99	0
724	Oleic Acid (9c-18:1) (% (w/w))	0164	0.7750	0.0500				2		724.99	0
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0164	1.895	0.0700				2		726.99	0
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0226	1.964	0.0451				2		726.99	0
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/w))	0226	0.1307	0.0029				2		728.99	0
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/w))	0164	0.1350	0.0100				2		728.99	0
730	Arachidic Acid (20:0) (% (w/w))	0226	0.0111	0.0000				1		730.99	0
730	Arachidic Acid (20:0) (% (w/w))	0164	< 0.02					1		730.99	5
732	Gondoic Acid (11c-20:1) (% (w/w))	0226	0.0180	0.0002				2		732.99	0
732	Gondoic Acid (11c-20:1) (% (w/w))	0164	0.0200	0.0000				2		732.99	0
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c) (% (w/w))	0226	< 0.005					0		740.99	5
742	Behenic Acid (22:0) (% (w/w))	0226	0.0089	0.0001				1		742.99	0
742	Behenic Acid (22:0) (% (w/w))	0164	< 0.02					1		742.99	5
744	Erucic Acid (13c-22:1) (% (w/w))	0226	0.0019	0.0001				1		744.99	0
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c) (% (w/w))	0226	< 0.005					0		746.99	5
748	Lignoceric Acid (24:0) (% (w/w))	0226	0.0125	0.0001				1		748.99	0
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22c) (% (w/w))	0226	< 0.005					0		750.99	5
752	Nervonic Acid (24:1) isomers (% (w/w))	0226	< 0.005					0		752.99	5
754	Total n-3 Polyunsaturated (Omega-3) Fatty Acids (% (w/w))	0164	0.1400	0.0200				1		754.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
756	Total n-6 Polyunsaturated (Omega-6) Fatty Acids	0164	1.885	0.0700				1	756.99	0	
758	Total Saturated Fatty Acids (% (w/w))	0164	0.8250	0.1100				1	758.99	0	
762	Total Monounsaturated Fatty Acids (% (w/w))	0164	0.8050	0.0500				1	762.99	0	
766	Total Polyunsaturated Fatty Acids (% (w/w))	0164	2.040	0.0800				1	766.99	0	
770	Total Fat (equivalent to NLEA) (% (w/w))	0164	3.860	0.2600				1	770.99	0	
772	Total Fatty Acids (% (w/w))	0226	3.547	0.0730				2	772.99	0	
772	Total Fatty Acids (% (w/w))	0164	3.695	0.2500				2	772.99	0	

Notes: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2 and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = Rejected for duplicates too far apart, 2 = Rejected as outlier, 5 = A reporting limit and 4 = zeros submitted as values. Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). Flag 3 indicates not used in statistics.