

Animal Feed Scheme
Catfish Feed
Labs Reporting: 191
Analytes Reported 109
Issue Date : 09/30/2019
Test Material Code # 201928
Analyte Proficiency Testing Report

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
000	Urea (%)	0278	0.2000	0.0000				1		000.02	0
001	Loss on Drying (%)	0618	5.450	0.4000	7.456	0.2774	0.0990	75	-7.23	001.99	0
001	Loss on Drying (%)	0536	6.305	0.1700	7.456	0.2774	0.0990	75	-4.15	001.99	0
001	Loss on Drying (%)	2155	6.340	0.1600	7.456	0.2774	0.0990	75	-4.02	001.99	0
001	Loss on Drying (%)	2062	6.363	0.0052	7.456	0.2774	0.0990	75	-3.94	001.03	0
001	Loss on Drying (%)	0038	6.530	0.0600	7.456	0.2774	0.0990	75	-3.34	001.07	0
001	Loss on Drying (%)	0028	6.530	0.1000	7.456	0.2774	0.0990	75	-3.34	001.07	0
001	Loss on Drying (%)	0889	6.670	0.1200	7.456	0.2774	0.0990	75	-2.83	001.99	0
001	Loss on Drying (%)	0353	6.845	0.1900	7.456	0.2774	0.0990	75	-2.20	001.07	0
001	Loss on Drying (%)	0918	6.985	0.0300	7.456	0.2774	0.0990	75	-1.70	001.99	0
001	Loss on Drying (%)	0505	6.985	0.0900	7.456	0.2774	0.0990	75	-1.70	001.07	0
001	Loss on Drying (%)	2262	7.018	0.0030	7.456	0.2774	0.0990	75	-1.58	001.07	0
001	Loss on Drying (%)	0510	7.050	0.1000	7.456	0.2774	0.0990	75	-1.46	001.99	0
001	Loss on Drying (%)	0504	7.115	0.0500	7.456	0.2774	0.0990	75	-1.23	001.00	0
001	Loss on Drying (%)	0226	7.145	0.1300	7.456	0.2774	0.0990	75	-1.12	001.07	0
001	Loss on Drying (%)	0853	7.200	0.2000	7.456	0.2774	0.0990	75	-0.92	001.07	0
001	Loss on Drying (%)	0610	7.220	0.0200	7.456	0.2774	0.0990	75	-0.85	001.05	0
001	Loss on Drying (%)	0034	7.235	0.1100	7.456	0.2774	0.0990	75	-0.80	001.07	0
001	Loss on Drying (%)	0868	7.255	0.0100	7.456	0.2774	0.0990	75	-0.72	001.03	0
001	Loss on Drying (%)	0083	7.315	0.0100	7.456	0.2774	0.0990	75	-0.51	001.07	0
001	Loss on Drying (%)	0142	7.325	0.0900	7.456	0.2774	0.0990	75	-0.47	001.07	0
001	Loss on Drying (%)	0171	7.350	0.0000	7.456	0.2774	0.0990	75	-0.38	001.07	0
001	Loss on Drying (%)	0849	7.355	0.0900	7.456	0.2774	0.0990	75	-0.36	001.07	0
001	Loss on Drying (%)	2192	7.365	0.1100	7.456	0.2774	0.0990	75	-0.33	001.07	0
001	Loss on Drying (%)	0278	7.400	0.2400	7.456	0.2774	0.0990	75	-0.20	001.07	0
001	Loss on Drying (%)	0186	7.418	0.0274	7.456	0.2774	0.0990	75	-0.14	001.99	0
001	Loss on Drying (%)	2009	7.428	0.0214	7.456	0.2774	0.0990	75	-0.10	001.07	0
001	Loss on Drying (%)	0035	7.440	0.0400	7.456	0.2774	0.0990	75	-0.06	001.07	0
001	Loss on Drying (%)	0682	7.440	0.0000	7.456	0.2774	0.0990	75	-0.06	001.07	0
001	Loss on Drying (%)	0074	7.460	0.0000	7.456	0.2774	0.0990	75	0.01	001.07	0
001	Loss on Drying (%)	2190	7.470	0.0000	7.456	0.2774	0.0990	75	0.05	001.07	0
001	Loss on Drying (%)	0644	7.470	0.0200	7.456	0.2774	0.0990	75	0.05	001.05	0
001	Loss on Drying (%)	0089	7.475	0.0300	7.456	0.2774	0.0990	75	0.07	001.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
001	Loss on Drying (%)	2259	7.489	0.1550	7.456	0.2774	0.0990	75	0.12	001.07	0
001	Loss on Drying (%)	0045	7.490	0.3400	7.456	0.2774	0.0990	75	0.12	001.07	0
001	Loss on Drying (%)	0098	7.490	0.1000	7.456	0.2774	0.0990	75	0.12	001.07	0
001	Loss on Drying (%)	0948	7.490	0.1000	7.456	0.2774	0.0990	75	0.12	001.99	0
001	Loss on Drying (%)	0872	7.495	0.0300	7.456	0.2774	0.0990	75	0.14	001.07	0
001	Loss on Drying (%)	0676	7.495	0.0700	7.456	0.2774	0.0990	75	0.14	001.99	0
001	Loss on Drying (%)	0693	7.499	0.0360	7.456	0.2774	0.0990	75	0.15	001.07	0
001	Loss on Drying (%)	0638	7.500	0.0400	7.456	0.2774	0.0990	75	0.16	001.07	0
001	Loss on Drying (%)	0169	7.510	0.0600	7.456	0.2774	0.0990	75	0.19	001.00	0
001	Loss on Drying (%)	2109	7.510	0.0000	7.456	0.2774	0.0990	75	0.19	001.07	0
001	Loss on Drying (%)	0571	7.525	0.0700	7.456	0.2774	0.0990	75	0.25	001.07	0
001	Loss on Drying (%)	2053	7.525	0.0500	7.456	0.2774	0.0990	75	0.25	001.07	0
001	Loss on Drying (%)	0720	7.525	0.2300	7.456	0.2774	0.0990	75	0.25	001.99	0
001	Loss on Drying (%)	0074	7.540	0.0000	7.456	0.2774	0.0990	75	0.30	001.05	0
001	Loss on Drying (%)	0037	7.545	0.0100	7.456	0.2774	0.0990	75	0.32	001.99	0
001	Loss on Drying (%)	2018	7.550	0.0350	7.456	0.2774	0.0990	75	0.34	001.99	0
001	Loss on Drying (%)	0148	7.550	0.0400	7.456	0.2774	0.0990	75	0.34	001.07	0
001	Loss on Drying (%)	0003	7.555	0.0500	7.456	0.2774	0.0990	75	0.36	001.07	0
001	Loss on Drying (%)	2012	7.555	0.0300	7.456	0.2774	0.0990	75	0.36	001.07	0
001	Loss on Drying (%)	2059	7.560	0.0600	7.456	0.2774	0.0990	75	0.37	001.07	0
001	Loss on Drying (%)	0683	7.565	0.0300	7.456	0.2774	0.0990	75	0.39	001.07	0
001	Loss on Drying (%)	0139	7.590	0.1000	7.456	0.2774	0.0990	75	0.48	001.07	0
001	Loss on Drying (%)	2146	7.630	0.2400	7.456	0.2774	0.0990	75	0.63	001.07	0
001	Loss on Drying (%)	0695	7.632	0.0140	7.456	0.2774	0.0990	75	0.63	001.07	0
001	Loss on Drying (%)	0675	7.640	0.0600	7.456	0.2774	0.0990	75	0.66	001.07	0
001	Loss on Drying (%)	0619	7.650	0.0600	7.456	0.2774	0.0990	75	0.70	001.03	0
001	Loss on Drying (%)	0066	7.650	0.1200	7.456	0.2774	0.0990	75	0.70	001.07	0
001	Loss on Drying (%)	0366	7.650	0.5000	7.456	0.2774	0.0990	75	0.70	001.07	0
001	Loss on Drying (%)	0689	7.650	0.1000	7.456	0.2774	0.0990	75	0.70	001.07	0
001	Loss on Drying (%)	0876	7.650	0.1000	7.456	0.2774	0.0990	75	0.70	001.07	0
001	Loss on Drying (%)	0843	7.665	0.1300	7.456	0.2774	0.0990	75	0.75	001.03	0
001	Loss on Drying (%)	0843	7.665	0.1300	7.456	0.2774	0.0990	75	0.75	001.07	0
001	Loss on Drying (%)	0049	7.675	0.0100	7.456	0.2774	0.0990	75	0.79	001.07	0
001	Loss on Drying (%)	0027	7.720	0.0710	7.456	0.2774	0.0990	75	0.95	001.00	0
001	Loss on Drying (%)	2268	7.720	0.3000	7.456	0.2774	0.0990	75	0.95	001.99	0
001	Loss on Drying (%)	0581	7.725	0.0500	7.456	0.2774	0.0990	75	0.97	001.07	0
001	Loss on Drying (%)	0199	7.750	0.1200	7.456	0.2774	0.0990	75	1.06	001.07	0
001	Loss on Drying (%)	0910	7.850	0.5000	7.456	0.2774	0.0990	75	1.42	001.99	0
001	Loss on Drying (%)	0590	7.900	0.3400	7.456	0.2774	0.0990	75	1.60	001.07	0
001	Loss on Drying (%)	0357	7.936	0.0771	7.456	0.2774	0.0990	75	1.73	001.99	0
001	Loss on Drying (%)	2246	7.950	0.1800	7.456	0.2774	0.0990	75	1.78	001.99	0
001	Loss on Drying (%)	0852	7.995	0.0100	7.456	0.2774	0.0990	75	1.94	001.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
001	Loss on Drying (%)	0529	8.365	0.0500	7.456	0.2774	0.0990	75	3.28	001.07	0
001	Loss on Drying (%)	2081	7.825	0.8500	7.456	0.2774	0.0990	75	1.33	001.99	1
001	Loss on Drying (%)	0560	7.900	1.000	7.456	0.2774	0.0990	75	1.60	001.99	1
002	Protein, Crude (%)	0852	31.39	0.0400	32.87	0.3304	0.1756	189	-4.48	002.11	0
002	Protein, Crude (%)	0644	31.47	0.1310	32.87	0.3304	0.1756	189	-4.24	002.06	0
002	Protein, Crude (%)	0874	31.75	0.2900	32.87	0.3304	0.1756	189	-3.41	002.04	0
002	Protein, Crude (%)	0553	31.87	0.0300	32.87	0.3304	0.1756	189	-3.04	002.11	0
002	Protein, Crude (%)	0720	31.94	0.0300	32.87	0.3304	0.1756	189	-2.83	002.11	0
002	Protein, Crude (%)	0720	32.00	0.1200	32.87	0.3304	0.1756	189	-2.63	002.06	0
002	Protein, Crude (%)	0885	32.01	0.0600	32.87	0.3304	0.1756	189	-2.62	002.05	0
002	Protein, Crude (%)	2248	32.02	0.1400	32.87	0.3304	0.1756	189	-2.57	002.05	0
002	Protein, Crude (%)	2190	32.06	0.4100	32.87	0.3304	0.1756	189	-2.47	002.05	0
002	Protein, Crude (%)	0650	32.06	0.1700	32.87	0.3304	0.1756	189	-2.47	002.06	0
002	Protein, Crude (%)	0910	32.10	0.0000	32.87	0.3304	0.1756	189	-2.33	002.01	0
002	Protein, Crude (%)	0598	32.18	0.0500	32.87	0.3304	0.1756	189	-2.10	002.06	0
002	Protein, Crude (%)	0689	32.25	0.1000	32.87	0.3304	0.1756	189	-1.88	002.05	0
002	Protein, Crude (%)	2006	32.33	0.1400	32.87	0.3304	0.1756	189	-1.64	002.05	0
002	Protein, Crude (%)	2058	32.38	0.0300	32.87	0.3304	0.1756	189	-1.50	002.09	0
002	Protein, Crude (%)	0955	32.40	0.2000	32.87	0.3304	0.1756	189	-1.42	002.05	0
002	Protein, Crude (%)	0510	32.40	0.2000	32.87	0.3304	0.1756	189	-1.42	002.06	0
002	Protein, Crude (%)	2198	32.41	0.0400	32.87	0.3304	0.1756	189	-1.39	002.01	0
002	Protein, Crude (%)	2073	32.43	0.2400	32.87	0.3304	0.1756	189	-1.33	002.05	0
002	Protein, Crude (%)	2109	32.43	0.1400	32.87	0.3304	0.1756	189	-1.33	002.05	0
002	Protein, Crude (%)	0539	32.43	0.0000	32.87	0.3304	0.1756	189	-1.33	002.06	0
002	Protein, Crude (%)	0651	32.43	0.1040	32.87	0.3304	0.1756	189	-1.32	002.05	0
002	Protein, Crude (%)	0536	32.44	0.1100	32.87	0.3304	0.1756	189	-1.32	002.05	0
002	Protein, Crude (%)	0592	32.45	0.1300	32.87	0.3304	0.1756	189	-1.29	002.06	0
002	Protein, Crude (%)	0563	32.45	0.0909	32.87	0.3304	0.1756	189	-1.26	002.08	0
002	Protein, Crude (%)	0675	32.46	0.0600	32.87	0.3304	0.1756	189	-1.24	002.05	0
002	Protein, Crude (%)	0638	32.47	0.2600	32.87	0.3304	0.1756	189	-1.21	002.04	0
002	Protein, Crude (%)	2251	32.48	0.3000	32.87	0.3304	0.1756	189	-1.18	002.06	0
002	Protein, Crude (%)	2062	32.50	0.0491	32.87	0.3304	0.1756	189	-1.12	002.05	0
002	Protein, Crude (%)	2146	32.50	0.1800	32.87	0.3304	0.1756	189	-1.12	002.05	0
002	Protein, Crude (%)	0955	32.50	0.4000	32.87	0.3304	0.1756	189	-1.12	002.06	0
002	Protein, Crude (%)	0948	32.54	0.0800	32.87	0.3304	0.1756	189	-1.00	002.05	0
002	Protein, Crude (%)	0354	32.56	0.1400	32.87	0.3304	0.1756	189	-0.94	002.05	0
002	Protein, Crude (%)	0881	32.57	0.0660	32.87	0.3304	0.1756	189	-0.91	002.05	0
002	Protein, Crude (%)	2199	32.60	0.0000	32.87	0.3304	0.1756	189	-0.82	002.05	0
002	Protein, Crude (%)	0870	32.61	0.2560	32.87	0.3304	0.1756	189	-0.79	002.01	0
002	Protein, Crude (%)	0169	32.62	0.1500	32.87	0.3304	0.1756	189	-0.77	002.02	0
002	Protein, Crude (%)	0529	32.62	0.1300	32.87	0.3304	0.1756	189	-0.77	002.06	0
002	Protein, Crude (%)	0964	32.62	0.4620	32.87	0.3304	0.1756	189	-0.76	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein, Crude (%)	0164	32.63	0.2500	32.87	0.3304	0.1756	189	-0.74	002.01	0
002	Protein, Crude (%)	2075	32.63	0.0170	32.87	0.3304	0.1756	189	-0.74	002.06	0
002	Protein, Crude (%)	0968	32.63	0.0680	32.87	0.3304	0.1756	189	-0.74	002.01	0
002	Protein, Crude (%)	0849	32.63	0.0420	32.87	0.3304	0.1756	189	-0.73	002.05	0
002	Protein, Crude (%)	2012	32.64	0.1400	32.87	0.3304	0.1756	189	-0.70	002.05	0
002	Protein, Crude (%)	0226	32.65	0.1000	32.87	0.3304	0.1756	189	-0.67	002.06	0
002	Protein, Crude (%)	2129	32.65	0.1000	32.87	0.3304	0.1756	189	-0.67	002.06	0
002	Protein, Crude (%)	2009	32.65	0.0399	32.87	0.3304	0.1756	189	-0.65	002.05	0
002	Protein, Crude (%)	0504	32.66	0.7700	32.87	0.3304	0.1756	189	-0.65	002.04	0
002	Protein, Crude (%)	0357	32.66	0.0625	32.87	0.3304	0.1756	189	-0.65	002.06	0
002	Protein, Crude (%)	0036	32.67	0.2890	32.87	0.3304	0.1756	189	-0.62	002.02	0
002	Protein, Crude (%)	2193	32.69	0.5000	32.87	0.3304	0.1756	189	-0.55	002.01	0
002	Protein, Crude (%)	0897	32.70	0.0300	32.87	0.3304	0.1756	189	-0.53	002.05	0
002	Protein, Crude (%)	0119	32.70	1.073	32.87	0.3304	0.1756	189	-0.53	002.06	0
002	Protein, Crude (%)	0626	32.70	0.0190	32.87	0.3304	0.1756	189	-0.52	002.06	0
002	Protein, Crude (%)	0652	32.70	0.2000	32.87	0.3304	0.1756	189	-0.52	002.01	0
002	Protein, Crude (%)	0100	32.70	0.1200	32.87	0.3304	0.1756	189	-0.52	002.06	0
002	Protein, Crude (%)	0062	32.71	0.1120	32.87	0.3304	0.1756	189	-0.49	002.06	0
002	Protein, Crude (%)	0139	32.71	0.1000	32.87	0.3304	0.1756	189	-0.49	002.06	0
002	Protein, Crude (%)	0098	32.71	0.0000	32.87	0.3304	0.1756	189	-0.49	002.08	0
002	Protein, Crude (%)	2150	32.73	0.0700	32.87	0.3304	0.1756	189	-0.44	002.06	0
002	Protein, Crude (%)	0066	32.73	0.1100	32.87	0.3304	0.1756	189	-0.44	002.02	0
002	Protein, Crude (%)	0868	32.73	0.0200	32.87	0.3304	0.1756	189	-0.43	002.05	0
002	Protein, Crude (%)	0186	32.74	0.7900	32.87	0.3304	0.1756	189	-0.41	002.06	0
002	Protein, Crude (%)	2022	32.74	0.1900	32.87	0.3304	0.1756	189	-0.41	002.06	0
002	Protein, Crude (%)	0089	32.75	0.0100	32.87	0.3304	0.1756	189	-0.38	002.06	0
002	Protein, Crude (%)	0619	32.75	0.3000	32.87	0.3304	0.1756	189	-0.36	002.05	0
002	Protein, Crude (%)	2022	32.75	0.0400	32.87	0.3304	0.1756	189	-0.36	002.05	0
002	Protein, Crude (%)	0889	32.77	0.1900	32.87	0.3304	0.1756	189	-0.32	002.01	0
002	Protein, Crude (%)	0675	32.79	0.0500	32.87	0.3304	0.1756	189	-0.26	002.06	0
002	Protein, Crude (%)	2128	32.79	0.0000	32.87	0.3304	0.1756	189	-0.26	002.03	0
002	Protein, Crude (%)	0171	32.79	0.0800	32.87	0.3304	0.1756	189	-0.24	002.06	0
002	Protein, Crude (%)	0682	32.79	0.0000	32.87	0.3304	0.1756	189	-0.24	002.06	0
002	Protein, Crude (%)	2188	32.80	0.0000	32.87	0.3304	0.1756	189	-0.21	002.01	0
002	Protein, Crude (%)	2097	32.80	0.1450	32.87	0.3304	0.1756	189	-0.21	002.05	0
002	Protein, Crude (%)	0047	32.81	0.7100	32.87	0.3304	0.1756	189	-0.20	002.06	0
002	Protein, Crude (%)	0674	32.81	0.0000	32.87	0.3304	0.1756	189	-0.18	002.05	0
002	Protein, Crude (%)	0683	32.81	0.4000	32.87	0.3304	0.1756	189	-0.18	002.05	0
002	Protein, Crude (%)	0425	32.81	0.2800	32.87	0.3304	0.1756	189	-0.18	002.06	0
002	Protein, Crude (%)	0202	32.82	0.1190	32.87	0.3304	0.1756	189	-0.17	002.06	0
002	Protein, Crude (%)	0939	32.82	0.4000	32.87	0.3304	0.1756	189	-0.15	002.05	0
002	Protein, Crude (%)	0868	32.83	0.0100	32.87	0.3304	0.1756	189	-0.14	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein, Crude (%)	2018	32.83	0.0800	32.87	0.3304	0.1756	189	-0.12	002.06	0
002	Protein, Crude (%)	0695	32.84	0.6860	32.87	0.3304	0.1756	189	-0.08	002.06	0
002	Protein, Crude (%)	0032	32.84	0.1390	32.87	0.3304	0.1756	189	-0.08	002.06	0
002	Protein, Crude (%)	0581	32.85	1.130	32.87	0.3304	0.1756	189	-0.08	002.06	0
002	Protein, Crude (%)	0278	32.85	0.1000	32.87	0.3304	0.1756	189	-0.06	002.06	0
002	Protein, Crude (%)	0520	32.85	0.1000	32.87	0.3304	0.1756	189	-0.06	002.06	0
002	Protein, Crude (%)	0610	32.85	0.1000	32.87	0.3304	0.1756	189	-0.06	002.06	0
002	Protein, Crude (%)	2196	32.85	0.0000	32.87	0.3304	0.1756	189	-0.06	002.01	0
002	Protein, Crude (%)	0884	32.85	0.3000	32.87	0.3304	0.1756	189	-0.06	002.06	0
002	Protein, Crude (%)	0074	32.86	0.7000	32.87	0.3304	0.1756	189	-0.03	002.06	0
002	Protein, Crude (%)	0358	32.87	0.2700	32.87	0.3304	0.1756	189	-0.02	002.06	0
002	Protein, Crude (%)	2246	32.88	0.0300	32.87	0.3304	0.1756	189	0.01	002.05	0
002	Protein, Crude (%)	0038	32.88	0.0800	32.87	0.3304	0.1756	189	0.03	002.06	0
002	Protein, Crude (%)	0968	32.89	0.1250	32.87	0.3304	0.1756	189	0.05	002.06	0
002	Protein, Crude (%)	0199	32.89	0.0000	32.87	0.3304	0.1756	189	0.06	002.01	0
002	Protein, Crude (%)	0619	32.90	0.0000	32.87	0.3304	0.1756	189	0.09	002.06	0
002	Protein, Crude (%)	0687	32.90	0.2000	32.87	0.3304	0.1756	189	0.09	002.06	0
002	Protein, Crude (%)	2152	32.90	0.1000	32.87	0.3304	0.1756	189	0.09	002.05	0
002	Protein, Crude (%)	2184	32.91	0.0800	32.87	0.3304	0.1756	189	0.12	002.06	0
002	Protein, Crude (%)	0036	32.92	0.2300	32.87	0.3304	0.1756	189	0.13	002.06	0
002	Protein, Crude (%)	0042	32.92	0.1100	32.87	0.3304	0.1756	189	0.13	002.06	0
002	Protein, Crude (%)	0108	32.92	0.0300	32.87	0.3304	0.1756	189	0.13	002.06	0
002	Protein, Crude (%)	0354	32.92	0.0200	32.87	0.3304	0.1756	189	0.15	002.06	0
002	Protein, Crude (%)	0298	32.93	0.0200	32.87	0.3304	0.1756	189	0.18	002.06	0
002	Protein, Crude (%)	0190	32.94	0.0900	32.87	0.3304	0.1756	189	0.20	002.06	0
002	Protein, Crude (%)	0018	32.94	0.2600	32.87	0.3304	0.1756	189	0.21	002.06	0
002	Protein, Crude (%)	0353	32.94	0.1000	32.87	0.3304	0.1756	189	0.21	002.06	0
002	Protein, Crude (%)	0848	32.94	0.0400	32.87	0.3304	0.1756	189	0.21	002.06	0
002	Protein, Crude (%)	2259	32.94	0.0270	32.87	0.3304	0.1756	189	0.22	002.01	0
002	Protein, Crude (%)	0098	32.95	0.1000	32.87	0.3304	0.1756	189	0.24	002.06	0
002	Protein, Crude (%)	0574	32.95	0.1000	32.87	0.3304	0.1756	189	0.24	002.06	0
002	Protein, Crude (%)	0546	32.95	0.0400	32.87	0.3304	0.1756	189	0.25	002.06	0
002	Protein, Crude (%)	0541	32.96	0.3900	32.87	0.3304	0.1756	189	0.26	002.06	0
002	Protein, Crude (%)	0027	32.96	0.0870	32.87	0.3304	0.1756	189	0.26	002.06	0
002	Protein, Crude (%)	0870	32.97	0.0325	32.87	0.3304	0.1756	189	0.30	002.06	0
002	Protein, Crude (%)	0726	32.98	0.0100	32.87	0.3304	0.1756	189	0.32	002.06	0
002	Protein, Crude (%)	0009	32.98	0.3900	32.87	0.3304	0.1756	189	0.32	002.06	0
002	Protein, Crude (%)	0693	32.98	0.1740	32.87	0.3304	0.1756	189	0.33	002.06	0
002	Protein, Crude (%)	0939	32.99	0.1300	32.87	0.3304	0.1756	189	0.35	002.06	0
002	Protein, Crude (%)	0144	33.00	0.0700	32.87	0.3304	0.1756	189	0.38	002.06	0
002	Protein, Crude (%)	2168	33.00	0.1500	32.87	0.3304	0.1756	189	0.38	002.06	0
002	Protein, Crude (%)	2155	33.00	0.2240	32.87	0.3304	0.1756	189	0.39	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein, Crude (%)	0164	33.00	0.1200	32.87	0.3304	0.1756	189	0.39	002.06	0
002	Protein, Crude (%)	0175	33.00	0.0000	32.87	0.3304	0.1756	189	0.39	002.06	0
002	Protein, Crude (%)	0366	33.00	0.0000	32.87	0.3304	0.1756	189	0.39	002.06	0
002	Protein, Crude (%)	0880	33.00	0.2000	32.87	0.3304	0.1756	189	0.39	002.06	0
002	Protein, Crude (%)	0049	33.01	0.5600	32.87	0.3304	0.1756	189	0.42	002.06	0
002	Protein, Crude (%)	0001	33.02	0.2500	32.87	0.3304	0.1756	189	0.44	002.06	0
002	Protein, Crude (%)	0006	33.03	0.0500	32.87	0.3304	0.1756	189	0.47	002.06	0
002	Protein, Crude (%)	0857	33.03	0.1200	32.87	0.3304	0.1756	189	0.48	002.06	0
002	Protein, Crude (%)	0589	33.05	0.2500	32.87	0.3304	0.1756	189	0.53	002.06	0
002	Protein, Crude (%)	0610	33.05	0.1000	32.87	0.3304	0.1756	189	0.54	002.01	0
002	Protein, Crude (%)	0010	33.05	0.3000	32.87	0.3304	0.1756	189	0.54	002.06	0
002	Protein, Crude (%)	0017	33.05	0.1000	32.87	0.3304	0.1756	189	0.54	002.06	0
002	Protein, Crude (%)	0045	33.05	0.1000	32.87	0.3304	0.1756	189	0.54	002.06	0
002	Protein, Crude (%)	0590	33.05	0.1400	32.87	0.3304	0.1756	189	0.54	002.06	0
002	Protein, Crude (%)	0910	33.05	0.1000	32.87	0.3304	0.1756	189	0.54	002.06	0
002	Protein, Crude (%)	0953	33.06	0.3100	32.87	0.3304	0.1756	189	0.56	002.06	0
002	Protein, Crude (%)	0142	33.06	0.1200	32.87	0.3304	0.1756	189	0.57	002.06	0
002	Protein, Crude (%)	0508	33.06	0.1310	32.87	0.3304	0.1756	189	0.58	002.06	0
002	Protein, Crude (%)	0918	33.08	0.0700	32.87	0.3304	0.1756	189	0.62	002.06	0
002	Protein, Crude (%)	0242	33.08	0.1000	32.87	0.3304	0.1756	189	0.63	002.06	0
002	Protein, Crude (%)	0148	33.10	0.2000	32.87	0.3304	0.1756	189	0.69	002.06	0
002	Protein, Crude (%)	0026	33.10	0.0570	32.87	0.3304	0.1756	189	0.70	002.06	0
002	Protein, Crude (%)	0553	33.11	0.1100	32.87	0.3304	0.1756	189	0.71	002.06	0
002	Protein, Crude (%)	0035	33.11	0.1800	32.87	0.3304	0.1756	189	0.72	002.06	0
002	Protein, Crude (%)	2192	33.12	0.2200	32.87	0.3304	0.1756	189	0.76	002.05	0
002	Protein, Crude (%)	0622	33.12	0.2265	32.87	0.3304	0.1756	189	0.76	002.05	0
002	Protein, Crude (%)	0512	33.13	0.6100	32.87	0.3304	0.1756	189	0.77	002.06	0
002	Protein, Crude (%)	0646	33.13	0.0900	32.87	0.3304	0.1756	189	0.77	002.06	0
002	Protein, Crude (%)	0948	33.13	0.0500	32.87	0.3304	0.1756	189	0.77	002.06	0
002	Protein, Crude (%)	2246	33.13	0.1300	32.87	0.3304	0.1756	189	0.77	002.06	0
002	Protein, Crude (%)	0504	33.14	0.2800	32.87	0.3304	0.1756	189	0.82	002.06	0
002	Protein, Crude (%)	0505	33.14	0.1800	32.87	0.3304	0.1756	189	0.82	002.06	0
002	Protein, Crude (%)	0083	33.15	0.2900	32.87	0.3304	0.1756	189	0.83	002.06	0
002	Protein, Crude (%)	2033	33.15	0.2600	32.87	0.3304	0.1756	189	0.85	002.06	0
002	Protein, Crude (%)	0037	33.17	0.4900	32.87	0.3304	0.1756	189	0.89	002.06	0
002	Protein, Crude (%)	2109	33.18	0.1200	32.87	0.3304	0.1756	189	0.94	002.06	0
002	Protein, Crude (%)	0843	33.19	0.4500	32.87	0.3304	0.1756	189	0.95	002.06	0
002	Protein, Crude (%)	2089	33.20	0.0100	32.87	0.3304	0.1756	189	0.98	002.06	0
002	Protein, Crude (%)	0294	33.20	0.0000	32.87	0.3304	0.1756	189	1.00	002.06	0
002	Protein, Crude (%)	0956	33.20	0.0000	32.87	0.3304	0.1756	189	1.00	002.06	0
002	Protein, Crude (%)	0003	33.21	0.0000	32.87	0.3304	0.1756	189	1.03	002.06	0
002	Protein, Crude (%)	0859	33.23	0.0770	32.87	0.3304	0.1756	189	1.08	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein, Crude (%)	0571	33.24	0.0000	32.87	0.3304	0.1756	189	1.11	002.06	0
002	Protein, Crude (%)	2081	33.24	0.6200	32.87	0.3304	0.1756	189	1.12	002.06	0
002	Protein, Crude (%)	0229	33.25	0.2700	32.87	0.3304	0.1756	189	1.13	002.06	0
002	Protein, Crude (%)	0676	33.26	0.2500	32.87	0.3304	0.1756	189	1.16	002.06	0
002	Protein, Crude (%)	0004	33.26	0.1800	32.87	0.3304	0.1756	189	1.18	002.06	0
002	Protein, Crude (%)	0872	33.29	0.2310	32.87	0.3304	0.1756	189	1.27	002.06	0
002	Protein, Crude (%)	0265	33.30	0.4000	32.87	0.3304	0.1756	189	1.30	002.06	0
002	Protein, Crude (%)	0876	33.35	0.1000	32.87	0.3304	0.1756	189	1.45	002.06	0
002	Protein, Crude (%)	0034	33.38	0.0390	32.87	0.3304	0.1756	189	1.54	002.06	0
002	Protein, Crude (%)	0199	33.45	0.1000	32.87	0.3304	0.1756	189	1.75	002.06	0
002	Protein, Crude (%)	0685	33.46	0.4700	32.87	0.3304	0.1756	189	1.77	002.01	0
002	Protein, Crude (%)	0263	33.48	0.1200	32.87	0.3304	0.1756	189	1.84	002.06	0
002	Protein, Crude (%)	2128	33.50	0.0690	32.87	0.3304	0.1756	189	1.90	002.11	0
002	Protein, Crude (%)	0011	33.50	0.0000	32.87	0.3304	0.1756	189	1.91	002.11	0
002	Protein, Crude (%)	2213	33.55	0.0500	32.87	0.3304	0.1756	189	2.04	002.06	0
002	Protein, Crude (%)	2059	33.56	0.0906	32.87	0.3304	0.1756	189	2.10	002.06	0
002	Protein, Crude (%)	2076	33.57	0.3110	32.87	0.3304	0.1756	189	2.10	002.06	0
002	Protein, Crude (%)	0407	33.61	0.9940	32.87	0.3304	0.1756	189	2.22	002.06	0
002	Protein, Crude (%)	0618	33.80	0.1260	32.87	0.3304	0.1756	189	2.82	002.06	0
002	Protein, Crude (%)	2262	33.81	0.0600	32.87	0.3304	0.1756	189	2.84	002.05	0
002	Protein, Crude (%)	0536	34.61	0.2100	32.87	0.3304	0.1756	189	5.25	002.03	0
002	Protein, Crude (%)	0853	35.20	0.6000	32.87	0.3304	0.1756	189	7.05	002.06	0
002	Protein, Crude (%)	0852	35.55	0.1800	32.87	0.3304	0.1756	189	8.11	002.05	0
002	Protein, Crude (%)	0405	35.72	0.0500	32.87	0.3304	0.1756	189	8.61	002.11	0
002	Protein, Crude (%)	0511	32.58	1.510	32.87	0.3304	0.1756	189	-0.89	002.06	1
002	Protein, Crude (%)	2053	32.65	1.700	32.87	0.3304	0.1756	189	-0.67	002.06	1
002	Protein, Crude (%)	2012	32.89	1.880	32.87	0.3304	0.1756	189	0.06	002.06	1
002	Protein, Crude (%)	2268	36.07	2.590	32.87	0.3304	0.1756	189	9.67	002.06	1
002	Protein, Crude (%)	0889	28.65	0.0400	32.87	0.3304	0.1756	189	-12.77	002.11	2
003	Fat, Crude (%)	0880	3.200	0.4000	4.101	0.2663	0.1021	142	-3.38	003.99	0
003	Fat, Crude (%)	0171	3.370	0.0000	4.101	0.2663	0.1021	142	-2.75	003.12	0
003	Fat, Crude (%)	2022	3.400	0.0800	4.101	0.2663	0.1021	142	-2.63	003.10	0
003	Fat, Crude (%)	0108	3.490	0.1000	4.101	0.2663	0.1021	142	-2.30	003.14	0
003	Fat, Crude (%)	0265	3.515	0.0300	4.101	0.2663	0.1021	142	-2.20	003.14	0
003	Fat, Crude (%)	0693	3.541	0.4270	4.101	0.2663	0.1021	142	-2.11	003.10	0
003	Fat, Crude (%)	2128	3.561	0.1229	4.101	0.2663	0.1021	142	-2.03	003.11	0
003	Fat, Crude (%)	0144	3.600	0.0600	4.101	0.2663	0.1021	142	-1.88	003.14	0
003	Fat, Crude (%)	2268	3.605	0.7100	4.101	0.2663	0.1021	142	-1.86	003.06	0
003	Fat, Crude (%)	2062	3.689	0.0023	4.101	0.2663	0.1021	142	-1.55	003.12	0
003	Fat, Crude (%)	0574	3.695	0.0300	4.101	0.2663	0.1021	142	-1.53	003.14	0
003	Fat, Crude (%)	2073	3.735	0.0100	4.101	0.2663	0.1021	142	-1.38	003.10	0
003	Fat, Crude (%)	0037	3.735	0.0300	4.101	0.2663	0.1021	142	-1.38	003.14	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat, Crude (%)	0853	3.750	0.1000	4.101	0.2663	0.1021	142	-1.32	003.14	0
003	Fat, Crude (%)	2155	3.760	0.0600	4.101	0.2663	0.1021	142	-1.28	003.10	0
003	Fat, Crude (%)	0553	3.770	0.0600	4.101	0.2663	0.1021	142	-1.24	003.10	0
003	Fat, Crude (%)	0581	3.770	0.1000	4.101	0.2663	0.1021	142	-1.24	003.14	0
003	Fat, Crude (%)	2184	3.780	0.0200	4.101	0.2663	0.1021	142	-1.21	003.10	0
003	Fat, Crude (%)	0870	3.795	0.2377	4.101	0.2663	0.1021	142	-1.15	003.10	0
003	Fat, Crude (%)	0357	3.800	0.0000	4.101	0.2663	0.1021	142	-1.13	003.14	0
003	Fat, Crude (%)	0876	3.800	0.4000	4.101	0.2663	0.1021	142	-1.13	003.14	0
003	Fat, Crude (%)	0695	3.815	0.2100	4.101	0.2663	0.1021	142	-1.07	003.10	0
003	Fat, Crude (%)	0948	3.820	0.0200	4.101	0.2663	0.1021	142	-1.06	003.14	0
003	Fat, Crude (%)	0202	3.835	0.0390	4.101	0.2663	0.1021	142	-1.00	003.14	0
003	Fat, Crude (%)	0062	3.845	0.0390	4.101	0.2663	0.1021	142	-0.96	003.10	0
003	Fat, Crude (%)	0036	3.882	0.0042	4.101	0.2663	0.1021	142	-0.82	003.10	0
003	Fat, Crude (%)	2198	3.885	0.1500	4.101	0.2663	0.1021	142	-0.81	003.00	0
003	Fat, Crude (%)	2193	3.885	0.0500	4.101	0.2663	0.1021	142	-0.81	003.10	0
003	Fat, Crude (%)	0186	3.895	0.0300	4.101	0.2663	0.1021	142	-0.77	003.14	0
003	Fat, Crude (%)	0066	3.910	0.0800	4.101	0.2663	0.1021	142	-0.72	003.14	0
003	Fat, Crude (%)	0098	3.925	0.0300	4.101	0.2663	0.1021	142	-0.66	003.10	0
003	Fat, Crude (%)	0263	3.928	0.0156	4.101	0.2663	0.1021	142	-0.65	003.10	0
003	Fat, Crude (%)	0674	3.950	0.0000	4.101	0.2663	0.1021	142	-0.57	003.09	0
003	Fat, Crude (%)	0948	3.955	0.1100	4.101	0.2663	0.1021	142	-0.55	003.09	0
003	Fat, Crude (%)	0100	3.965	0.0900	4.101	0.2663	0.1021	142	-0.51	003.10	0
003	Fat, Crude (%)	0619	3.965	0.0300	4.101	0.2663	0.1021	142	-0.51	003.10	0
003	Fat, Crude (%)	2129	3.965	0.0500	4.101	0.2663	0.1021	142	-0.51	003.14	0
003	Fat, Crude (%)	0881	3.970	0.1010	4.101	0.2663	0.1021	142	-0.49	003.10	0
003	Fat, Crude (%)	0298	3.970	0.0200	4.101	0.2663	0.1021	142	-0.49	003.14	0
003	Fat, Crude (%)	0619	3.975	0.0300	4.101	0.2663	0.1021	142	-0.47	003.14	0
003	Fat, Crude (%)	0683	3.975	0.2700	4.101	0.2663	0.1021	142	-0.47	003.14	0
003	Fat, Crude (%)	0045	3.980	0.0000	4.101	0.2663	0.1021	142	-0.46	003.10	0
003	Fat, Crude (%)	2196	3.980	0.0000	4.101	0.2663	0.1021	142	-0.46	003.10	0
003	Fat, Crude (%)	0511	3.980	0.0200	4.101	0.2663	0.1021	142	-0.46	003.14	0
003	Fat, Crude (%)	0541	3.995	0.1300	4.101	0.2663	0.1021	142	-0.40	003.10	0
003	Fat, Crude (%)	0098	3.995	0.2100	4.101	0.2663	0.1021	142	-0.40	003.13	0
003	Fat, Crude (%)	0003	3.995	0.4100	4.101	0.2663	0.1021	142	-0.40	003.14	0
003	Fat, Crude (%)	0358	3.995	0.2100	4.101	0.2663	0.1021	142	-0.40	003.14	0
003	Fat, Crude (%)	0689	4.000	0.0000	4.101	0.2663	0.1021	142	-0.38	003.06	0
003	Fat, Crude (%)	0689	4.000	0.0000	4.101	0.2663	0.1021	142	-0.38	003.10	0
003	Fat, Crude (%)	0089	4.015	0.0300	4.101	0.2663	0.1021	142	-0.32	003.14	0
003	Fat, Crude (%)	0651	4.016	0.0270	4.101	0.2663	0.1021	142	-0.32	003.10	0
003	Fat, Crude (%)	0638	4.025	0.0500	4.101	0.2663	0.1021	142	-0.29	003.09	0
003	Fat, Crude (%)	0169	4.030	0.0600	4.101	0.2663	0.1021	142	-0.27	003.06	0
003	Fat, Crude (%)	0042	4.030	0.1000	4.101	0.2663	0.1021	142	-0.27	003.10	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat, Crude (%)	0897	4.040	0.0000	4.101	0.2663	0.1021	142	-0.23	003.10	0
003	Fat, Crude (%)	2089	4.050	0.1800	4.101	0.2663	0.1021	142	-0.19	003.10	0
003	Fat, Crude (%)	0028	4.050	0.0000	4.101	0.2663	0.1021	142	-0.19	003.13	0
003	Fat, Crude (%)	0726	4.050	0.0800	4.101	0.2663	0.1021	142	-0.19	003.14	0
003	Fat, Crude (%)	0868	4.050	0.0200	4.101	0.2663	0.1021	142	-0.19	003.14	0
003	Fat, Crude (%)	0939	4.050	0.0400	4.101	0.2663	0.1021	142	-0.19	003.06	0
003	Fat, Crude (%)	2246	4.055	0.0500	4.101	0.2663	0.1021	142	-0.17	003.13	0
003	Fat, Crude (%)	2033	4.065	0.0500	4.101	0.2663	0.1021	142	-0.14	003.06	0
003	Fat, Crude (%)	0868	4.065	0.0700	4.101	0.2663	0.1021	142	-0.14	003.10	0
003	Fat, Crude (%)	0027	4.067	0.0130	4.101	0.2663	0.1021	142	-0.13	003.09	0
003	Fat, Crude (%)	2190	4.070	0.1200	4.101	0.2663	0.1021	142	-0.12	003.00	0
003	Fat, Crude (%)	0720	4.070	0.0400	4.101	0.2663	0.1021	142	-0.12	003.10	0
003	Fat, Crude (%)	0675	4.080	0.0600	4.101	0.2663	0.1021	142	-0.08	003.09	0
003	Fat, Crude (%)	2009	4.091	0.0334	4.101	0.2663	0.1021	142	-0.04	003.10	0
003	Fat, Crude (%)	0425	4.095	0.0700	4.101	0.2663	0.1021	142	-0.02	003.06	0
003	Fat, Crude (%)	2006	4.095	0.0100	4.101	0.2663	0.1021	142	-0.02	003.13	0
003	Fat, Crude (%)	0083	4.095	0.0300	4.101	0.2663	0.1021	142	-0.02	003.14	0
003	Fat, Crude (%)	2251	4.100	0.0200	4.101	0.2663	0.1021	142	0.00	003.07	0
003	Fat, Crude (%)	0354	4.100	0.0200	4.101	0.2663	0.1021	142	0.00	003.09	0
003	Fat, Crude (%)	2076	4.107	0.0465	4.101	0.2663	0.1021	142	0.02	003.14	0
003	Fat, Crude (%)	0646	4.115	0.1700	4.101	0.2663	0.1021	142	0.05	003.13	0
003	Fat, Crude (%)	2192	4.125	0.0100	4.101	0.2663	0.1021	142	0.09	003.00	0
003	Fat, Crude (%)	0644	4.132	0.1100	4.101	0.2663	0.1021	142	0.12	003.06	0
003	Fat, Crude (%)	0955	4.135	0.0700	4.101	0.2663	0.1021	142	0.13	003.14	0
003	Fat, Crude (%)	0229	4.137	0.0070	4.101	0.2663	0.1021	142	0.13	003.14	0
003	Fat, Crude (%)	0598	4.140	0.0000	4.101	0.2663	0.1021	142	0.15	003.14	0
003	Fat, Crude (%)	0175	4.140	0.2000	4.101	0.2663	0.1021	142	0.15	003.00	0
003	Fat, Crude (%)	0366	4.150	0.1000	4.101	0.2663	0.1021	142	0.18	003.10	0
003	Fat, Crude (%)	0278	4.150	0.1000	4.101	0.2663	0.1021	142	0.18	003.14	0
003	Fat, Crude (%)	0563	4.151	0.0143	4.101	0.2663	0.1021	142	0.19	003.14	0
003	Fat, Crude (%)	0229	4.152	0.0370	4.101	0.2663	0.1021	142	0.19	003.06	0
003	Fat, Crude (%)	0849	4.155	0.0300	4.101	0.2663	0.1021	142	0.20	003.06	0
003	Fat, Crude (%)	0529	4.155	0.1300	4.101	0.2663	0.1021	142	0.20	003.14	0
003	Fat, Crude (%)	0870	4.158	0.0299	4.101	0.2663	0.1021	142	0.21	003.13	0
003	Fat, Crude (%)	0148	4.160	0.0200	4.101	0.2663	0.1021	142	0.22	003.10	0
003	Fat, Crude (%)	0852	4.160	0.0000	4.101	0.2663	0.1021	142	0.22	003.11	0
003	Fat, Crude (%)	0098	4.165	0.1100	4.101	0.2663	0.1021	142	0.24	003.09	0
003	Fat, Crude (%)	0164	4.185	0.1100	4.101	0.2663	0.1021	142	0.31	003.06	0
003	Fat, Crude (%)	0687	4.200	0.2000	4.101	0.2663	0.1021	142	0.37	003.06	0
003	Fat, Crude (%)	0956	4.200	0.2000	4.101	0.2663	0.1021	142	0.37	003.14	0
003	Fat, Crude (%)	2097	4.203	0.0014	4.101	0.2663	0.1021	142	0.38	003.06	0
003	Fat, Crude (%)	0407	4.210	0.0535	4.101	0.2663	0.1021	142	0.41	003.14	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat, Crude (%)	0918	4.210	0.0200	4.101	0.2663	0.1021	142	0.41	003.06	0
003	Fat, Crude (%)	0676	4.210	0.0200	4.101	0.2663	0.1021	142	0.41	003.10	0
003	Fat, Crude (%)	0294	4.210	0.0400	4.101	0.2663	0.1021	142	0.41	003.06	0
003	Fat, Crude (%)	0870	4.211	0.1054	4.101	0.2663	0.1021	142	0.41	003.09	0
003	Fat, Crude (%)	2018	4.215	0.0300	4.101	0.2663	0.1021	142	0.43	003.10	0
003	Fat, Crude (%)	0148	4.215	0.1500	4.101	0.2663	0.1021	142	0.43	003.14	0
003	Fat, Crude (%)	2109	4.225	0.5700	4.101	0.2663	0.1021	142	0.46	003.00	0
003	Fat, Crude (%)	0536	4.225	0.0300	4.101	0.2663	0.1021	142	0.46	003.99	0
003	Fat, Crude (%)	0142	4.235	0.0100	4.101	0.2663	0.1021	142	0.50	003.00	0
003	Fat, Crude (%)	0009	4.235	0.0900	4.101	0.2663	0.1021	142	0.50	003.14	0
003	Fat, Crude (%)	0682	4.240	0.0000	4.101	0.2663	0.1021	142	0.52	003.06	0
003	Fat, Crude (%)	0889	4.243	0.0092	4.101	0.2663	0.1021	142	0.53	003.11	0
003	Fat, Crude (%)	0968	4.244	0.1810	4.101	0.2663	0.1021	142	0.53	003.14	0
003	Fat, Crude (%)	0682	4.260	0.0000	4.101	0.2663	0.1021	142	0.60	003.14	0
003	Fat, Crude (%)	0910	4.290	0.3800	4.101	0.2663	0.1021	142	0.71	003.06	0
003	Fat, Crude (%)	0520	4.290	0.1200	4.101	0.2663	0.1021	142	0.71	003.14	0
003	Fat, Crude (%)	0001	4.296	0.0480	4.101	0.2663	0.1021	142	0.73	003.14	0
003	Fat, Crude (%)	0885	4.298	0.0597	4.101	0.2663	0.1021	142	0.74	003.00	0
003	Fat, Crude (%)	2053	4.300	0.0200	4.101	0.2663	0.1021	142	0.75	003.00	0
003	Fat, Crude (%)	0884	4.300	0.2000	4.101	0.2663	0.1021	142	0.75	003.06	0
003	Fat, Crude (%)	2012	4.320	0.0600	4.101	0.2663	0.1021	142	0.82	003.09	0
003	Fat, Crude (%)	0199	4.325	0.0500	4.101	0.2663	0.1021	142	0.84	003.06	0
003	Fat, Crude (%)	2188	4.375	0.3100	4.101	0.2663	0.1021	142	1.03	003.10	0
003	Fat, Crude (%)	0190	4.410	0.0400	4.101	0.2663	0.1021	142	1.16	003.14	0
003	Fat, Crude (%)	0590	4.440	0.0200	4.101	0.2663	0.1021	142	1.27	003.09	0
003	Fat, Crude (%)	0504	4.445	0.0900	4.101	0.2663	0.1021	142	1.29	003.01	0
003	Fat, Crude (%)	0685	4.470	0.5000	4.101	0.2663	0.1021	142	1.38	003.09	0
003	Fat, Crude (%)	2262	4.475	0.0500	4.101	0.2663	0.1021	142	1.40	003.10	0
003	Fat, Crude (%)	0226	4.485	0.2300	4.101	0.2663	0.1021	142	1.44	003.09	0
003	Fat, Crude (%)	0049	4.485	0.6100	4.101	0.2663	0.1021	142	1.44	003.14	0
003	Fat, Crude (%)	0175	4.500	0.0800	4.101	0.2663	0.1021	142	1.50	003.14	0
003	Fat, Crude (%)	2058	4.610	0.0600	4.101	0.2663	0.1021	142	1.91	003.12	0
003	Fat, Crude (%)	2248	4.640	0.1600	4.101	0.2663	0.1021	142	2.02	003.01	0
003	Fat, Crude (%)	0353	4.670	0.5200	4.101	0.2663	0.1021	142	2.14	003.09	0
003	Fat, Crude (%)	0618	4.783	0.0260	4.101	0.2663	0.1021	142	2.56	003.10	0
003	Fat, Crude (%)	2199	4.850	0.7000	4.101	0.2663	0.1021	142	2.81	003.12	0
003	Fat, Crude (%)	2256	5.055	0.0500	4.101	0.2663	0.1021	142	3.58	003.00	0
003	Fat, Crude (%)	0011	5.100	0.0000	4.101	0.2663	0.1021	142	3.75	003.11	0
003	Fat, Crude (%)	0546	5.180	0.0400	4.101	0.2663	0.1021	142	4.05	003.99	0
003	Fat, Crude (%)	0720	5.235	0.0500	4.101	0.2663	0.1021	142	4.26	003.11	0
003	Fat, Crude (%)	0553	5.455	0.0500	4.101	0.2663	0.1021	142	5.08	003.11	0
003	Fat, Crude (%)	0405	6.015	0.0100	4.101	0.2663	0.1021	142	7.19	003.11	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat, Crude (%)	0941	6.155	0.0900	4.101	0.2663	0.1021	142	7.71	003.99	0
003	Fat, Crude (%)	0510	6.350	0.3000	4.101	0.2663	0.1021	142	8.45	003.09	0
003	Fat, Crude (%)	0171	7.025	0.0700	4.101	0.2663	0.1021	142	10.98	003.14	0
003	Fat, Crude (%)	0889	3.010	1.840	4.101	0.2663	0.1021	142	-4.10	003.10	1
003	Fat, Crude (%)	0843	3.080	0.8200	4.101	0.2663	0.1021	142	-3.84	003.14	1
003	Fat, Crude (%)	0242	4.255	1.030	4.101	0.2663	0.1021	142	0.58	003.10	1
004	Fiber, Crude (%)	2128	2.630	0.0199	6.942	0.7246	0.2046	101	-5.95	004.11	0
004	Fiber, Crude (%)	2089	5.170	0.0200	6.942	0.7246	0.2046	101	-2.45	004.03	0
004	Fiber, Crude (%)	0202	5.384	0.3610	6.942	0.7246	0.2046	101	-2.15	004.07	0
004	Fiber, Crude (%)	0144	5.615	0.0700	6.942	0.7246	0.2046	101	-1.83	004.07	0
004	Fiber, Crude (%)	0265	5.655	0.1700	6.942	0.7246	0.2046	101	-1.78	004.07	0
004	Fiber, Crude (%)	0553	5.715	0.2100	6.942	0.7246	0.2046	101	-1.69	004.11	0
004	Fiber, Crude (%)	2146	5.720	0.4000	6.942	0.7246	0.2046	101	-1.69	004.99	0
004	Fiber, Crude (%)	0708	5.730	0.0600	6.942	0.7246	0.2046	101	-1.67	004.07	0
004	Fiber, Crude (%)	0100	5.885	0.7100	6.942	0.7246	0.2046	101	-1.46	004.07	0
004	Fiber, Crude (%)	0682	6.000	0.0000	6.942	0.7246	0.2046	101	-1.30	004.07	0
004	Fiber, Crude (%)	0510	6.150	0.3000	6.942	0.7246	0.2046	101	-1.09	004.07	0
004	Fiber, Crude (%)	0968	6.160	0.1400	6.942	0.7246	0.2046	101	-1.08	004.07	0
004	Fiber, Crude (%)	0675	6.170	0.1200	6.942	0.7246	0.2046	101	-1.07	004.07	0
004	Fiber, Crude (%)	0226	6.240	0.4000	6.942	0.7246	0.2046	101	-0.97	004.00	0
004	Fiber, Crude (%)	0956	6.300	0.2000	6.942	0.7246	0.2046	101	-0.89	004.07	0
004	Fiber, Crude (%)	0405	6.330	0.0200	6.942	0.7246	0.2046	101	-0.84	004.11	0
004	Fiber, Crude (%)	0169	6.350	0.3000	6.942	0.7246	0.2046	101	-0.82	004.00	0
004	Fiber, Crude (%)	2012	6.350	0.0000	6.942	0.7246	0.2046	101	-0.82	004.07	0
004	Fiber, Crude (%)	2009	6.376	0.0497	6.942	0.7246	0.2046	101	-0.78	004.07	0
004	Fiber, Crude (%)	0263	6.379	0.0638	6.942	0.7246	0.2046	101	-0.78	004.07	0
004	Fiber, Crude (%)	0505	6.380	0.0600	6.942	0.7246	0.2046	101	-0.78	004.07	0
004	Fiber, Crude (%)	0366	6.400	0.6000	6.942	0.7246	0.2046	101	-0.75	004.01	0
004	Fiber, Crude (%)	2188	6.400	0.0000	6.942	0.7246	0.2046	101	-0.75	004.06	0
004	Fiber, Crude (%)	0726	6.420	0.0000	6.942	0.7246	0.2046	101	-0.72	004.07	0
004	Fiber, Crude (%)	0098	6.445	0.5700	6.942	0.7246	0.2046	101	-0.69	004.07	0
004	Fiber, Crude (%)	0689	6.450	0.1000	6.942	0.7246	0.2046	101	-0.68	004.06	0
004	Fiber, Crude (%)	0190	6.450	0.1000	6.942	0.7246	0.2046	101	-0.68	004.07	0
004	Fiber, Crude (%)	2190	6.450	0.7600	6.942	0.7246	0.2046	101	-0.68	004.07	0
004	Fiber, Crude (%)	0675	6.455	0.4300	6.942	0.7246	0.2046	101	-0.67	004.06	0
004	Fiber, Crude (%)	0939	6.475	0.0700	6.942	0.7246	0.2046	101	-0.64	004.07	0
004	Fiber, Crude (%)	0849	6.480	0.1200	6.942	0.7246	0.2046	101	-0.64	004.07	0
004	Fiber, Crude (%)	0229	6.495	0.2500	6.942	0.7246	0.2046	101	-0.62	004.07	0
004	Fiber, Crude (%)	0695	6.510	0.3600	6.942	0.7246	0.2046	101	-0.60	004.00	0
004	Fiber, Crude (%)	0619	6.510	0.5200	6.942	0.7246	0.2046	101	-0.60	004.07	0
004	Fiber, Crude (%)	0563	6.533	0.1467	6.942	0.7246	0.2046	101	-0.56	004.07	0
004	Fiber, Crude (%)	0590	6.540	0.8000	6.942	0.7246	0.2046	101	-0.55	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber, Crude (%)	0644	6.594	0.1760	6.942	0.7246	0.2046	101	-0.48	004.07	0
004	Fiber, Crude (%)	0910	6.650	0.1000	6.942	0.7246	0.2046	101	-0.40	004.07	0
004	Fiber, Crude (%)	2062	6.674	0.1031	6.942	0.7246	0.2046	101	-0.37	004.06	0
004	Fiber, Crude (%)	2033	6.700	0.0000	6.942	0.7246	0.2046	101	-0.33	004.07	0
004	Fiber, Crude (%)	0028	6.750	0.1000	6.942	0.7246	0.2046	101	-0.26	004.07	0
004	Fiber, Crude (%)	0885	6.770	0.1971	6.942	0.7246	0.2046	101	-0.24	004.06	0
004	Fiber, Crude (%)	2109	6.770	0.0400	6.942	0.7246	0.2046	101	-0.24	004.07	0
004	Fiber, Crude (%)	0354	6.790	0.0000	6.942	0.7246	0.2046	101	-0.21	004.07	0
004	Fiber, Crude (%)	2196	6.790	0.0000	6.942	0.7246	0.2046	101	-0.21	004.07	0
004	Fiber, Crude (%)	2193	6.795	0.0500	6.942	0.7246	0.2046	101	-0.20	004.07	0
004	Fiber, Crude (%)	0354	6.800	0.0200	6.942	0.7246	0.2046	101	-0.20	004.00	0
004	Fiber, Crude (%)	0520	6.800	0.2000	6.942	0.7246	0.2046	101	-0.20	004.07	0
004	Fiber, Crude (%)	0098	6.830	0.2600	6.942	0.7246	0.2046	101	-0.15	004.06	0
004	Fiber, Crude (%)	0626	6.850	0.3800	6.942	0.7246	0.2046	101	-0.13	004.07	0
004	Fiber, Crude (%)	0683	6.855	0.0300	6.942	0.7246	0.2046	101	-0.12	004.07	0
004	Fiber, Crude (%)	0083	6.880	0.1400	6.942	0.7246	0.2046	101	-0.09	004.07	0
004	Fiber, Crude (%)	2006	6.895	0.5700	6.942	0.7246	0.2046	101	-0.06	004.06	0
004	Fiber, Crude (%)	0164	6.900	0.2000	6.942	0.7246	0.2046	101	-0.06	004.00	0
004	Fiber, Crude (%)	0175	6.950	0.1000	6.942	0.7246	0.2046	101	0.01	004.00	0
004	Fiber, Crude (%)	0353	7.045	0.1500	6.942	0.7246	0.2046	101	0.14	004.03	0
004	Fiber, Crude (%)	0870	7.060	0.2637	6.942	0.7246	0.2046	101	0.16	004.07	0
004	Fiber, Crude (%)	0294	7.100	0.0000	6.942	0.7246	0.2046	101	0.22	004.07	0
004	Fiber, Crude (%)	0407	7.122	0.2357	6.942	0.7246	0.2046	101	0.25	004.07	0
004	Fiber, Crude (%)	0868	7.190	0.0800	6.942	0.7246	0.2046	101	0.34	004.06	0
004	Fiber, Crude (%)	2188	7.195	0.0700	6.942	0.7246	0.2046	101	0.35	004.07	0
004	Fiber, Crude (%)	0298	7.200	0.2000	6.942	0.7246	0.2046	101	0.36	004.07	0
004	Fiber, Crude (%)	0425	7.200	0.0000	6.942	0.7246	0.2046	101	0.36	004.00	0
004	Fiber, Crude (%)	0638	7.200	0.4000	6.942	0.7246	0.2046	101	0.36	004.06	0
004	Fiber, Crude (%)	0504	7.255	0.4900	6.942	0.7246	0.2046	101	0.43	004.00	0
004	Fiber, Crude (%)	0674	7.260	0.0000	6.942	0.7246	0.2046	101	0.44	004.06	0
004	Fiber, Crude (%)	0089	7.275	0.0700	6.942	0.7246	0.2046	101	0.46	004.07	0
004	Fiber, Crude (%)	0300	7.305	0.0100	6.942	0.7246	0.2046	101	0.50	004.07	0
004	Fiber, Crude (%)	0598	7.335	0.0100	6.942	0.7246	0.2046	101	0.54	004.07	0
004	Fiber, Crude (%)	0881	7.350	0.5300	6.942	0.7246	0.2046	101	0.56	004.06	0
004	Fiber, Crude (%)	2022	7.390	0.1400	6.942	0.7246	0.2046	101	0.62	004.06	0
004	Fiber, Crude (%)	0199	7.445	0.0900	6.942	0.7246	0.2046	101	0.69	004.06	0
004	Fiber, Crude (%)	2076	7.458	0.0242	6.942	0.7246	0.2046	101	0.71	004.00	0
004	Fiber, Crude (%)	0066	7.495	0.2300	6.942	0.7246	0.2046	101	0.76	004.07	0
004	Fiber, Crude (%)	0646	7.540	0.2800	6.942	0.7246	0.2046	101	0.83	004.07	0
004	Fiber, Crude (%)	0897	7.545	0.0100	6.942	0.7246	0.2046	101	0.83	004.06	0
004	Fiber, Crude (%)	0884	7.550	0.1000	6.942	0.7246	0.2046	101	0.84	004.00	0
004	Fiber, Crude (%)	0876	7.550	0.9000	6.942	0.7246	0.2046	101	0.84	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
004	Fiber, Crude (%)	0948	7.565	0.2100	6.942	0.7246	0.2046	101	0.86	004.06	0
004	Fiber, Crude (%)	0968	7.573	0.0540	6.942	0.7246	0.2046	101	0.87	004.06	0
004	Fiber, Crude (%)	0918	7.575	0.8100	6.942	0.7246	0.2046	101	0.87	004.06	0
004	Fiber, Crude (%)	0720	7.580	0.0000	6.942	0.7246	0.2046	101	0.88	004.06	0
004	Fiber, Crude (%)	2259	7.655	0.0130	6.942	0.7246	0.2046	101	0.98	004.07	0
004	Fiber, Crude (%)	0553	7.670	0.1000	6.942	0.7246	0.2046	101	1.01	004.07	0
004	Fiber, Crude (%)	0693	7.700	0.2030	6.942	0.7246	0.2046	101	1.05	004.07	0
004	Fiber, Crude (%)	0948	7.710	0.3000	6.942	0.7246	0.2046	101	1.06	004.07	0
004	Fiber, Crude (%)	0581	7.720	0.1200	6.942	0.7246	0.2046	101	1.07	004.07	0
004	Fiber, Crude (%)	0042	7.735	0.4300	6.942	0.7246	0.2046	101	1.09	004.07	0
004	Fiber, Crude (%)	0889	7.747	0.4995	6.942	0.7246	0.2046	101	1.11	004.11	0
004	Fiber, Crude (%)	0045	7.760	0.4600	6.942	0.7246	0.2046	101	1.13	004.07	0
004	Fiber, Crude (%)	2198	7.775	0.1900	6.942	0.7246	0.2046	101	1.15	004.06	0
004	Fiber, Crude (%)	0034	7.775	0.4700	6.942	0.7246	0.2046	101	1.15	004.07	0
004	Fiber, Crude (%)	2058	7.790	0.1200	6.942	0.7246	0.2046	101	1.17	004.03	0
004	Fiber, Crude (%)	2073	7.855	0.0900	6.942	0.7246	0.2046	101	1.26	004.00	0
004	Fiber, Crude (%)	2192	7.855	0.3900	6.942	0.7246	0.2046	101	1.26	004.03	0
004	Fiber, Crude (%)	2246	7.910	0.2000	6.942	0.7246	0.2046	101	1.34	004.03	0
004	Fiber, Crude (%)	0278	8.300	0.0000	6.942	0.7246	0.2046	101	1.87	004.07	0
004	Fiber, Crude (%)	0643	8.605	0.2500	6.942	0.7246	0.2046	101	2.30	004.07	0
004	Fiber, Crude (%)	0536	8.855	0.0900	6.942	0.7246	0.2046	101	2.64	004.07	0
004	Fiber, Crude (%)	0720	9.700	0.2000	6.942	0.7246	0.2046	101	3.81	004.11	0
004	Fiber, Crude (%)	0171	10.07	0.0100	6.942	0.7246	0.2046	101	4.31	004.00	0
005	Ash (%)	0720	3.365	0.0100	6.254	0.0953	0.0520	141	-30.32	005.11	0
005	Ash (%)	0643	5.730	0.1000	6.254	0.0953	0.0520	141	-5.50	005.00	0
005	Ash (%)	0169	5.965	0.1100	6.254	0.0953	0.0520	141	-3.04	005.00	0
005	Ash (%)	2184	5.995	0.0500	6.254	0.0953	0.0520	141	-2.72	005.00	0
005	Ash (%)	0953	6.010	0.0200	6.254	0.0953	0.0520	141	-2.56	005.00	0
005	Ash (%)	0853	6.050	0.1000	6.254	0.0953	0.0520	141	-2.14	005.00	0
005	Ash (%)	2199	6.050	0.1000	6.254	0.0953	0.0520	141	-2.14	005.00	0
005	Ash (%)	0880	6.070	0.1800	6.254	0.0953	0.0520	141	-1.93	005.00	0
005	Ash (%)	0646	6.100	0.0200	6.254	0.0953	0.0520	141	-1.62	005.00	0
005	Ash (%)	2033	6.100	0.0000	6.254	0.0953	0.0520	141	-1.62	005.00	0
005	Ash (%)	0229	6.110	0.0200	6.254	0.0953	0.0520	141	-1.51	005.00	0
005	Ash (%)	0358	6.115	0.0500	6.254	0.0953	0.0520	141	-1.46	005.00	0
005	Ash (%)	0186	6.115	0.0456	6.254	0.0953	0.0520	141	-1.46	005.00	0
005	Ash (%)	0049	6.125	0.0300	6.254	0.0953	0.0520	141	-1.36	005.00	0
005	Ash (%)	2256	6.130	0.0400	6.254	0.0953	0.0520	141	-1.31	005.00	0
005	Ash (%)	0541	6.135	0.1100	6.254	0.0953	0.0520	141	-1.25	005.00	0
005	Ash (%)	0539	6.140	0.0400	6.254	0.0953	0.0520	141	-1.20	005.00	0
005	Ash (%)	0574	6.150	0.1000	6.254	0.0953	0.0520	141	-1.10	005.00	0
005	Ash (%)	2006	6.150	0.0400	6.254	0.0953	0.0520	141	-1.10	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0618	6.153	0.0301	6.254	0.0953	0.0520	141	-1.07	005.00	0
005	Ash (%)	0175	6.155	0.0100	6.254	0.0953	0.0520	141	-1.04	005.00	0
005	Ash (%)	0004	6.160	0.0200	6.254	0.0953	0.0520	141	-0.99	005.00	0
005	Ash (%)	0278	6.160	0.0600	6.254	0.0953	0.0520	141	-0.99	005.00	0
005	Ash (%)	0407	6.166	0.0107	6.254	0.0953	0.0520	141	-0.93	005.00	0
005	Ash (%)	0144	6.170	0.0200	6.254	0.0953	0.0520	141	-0.89	005.00	0
005	Ash (%)	0918	6.175	0.0700	6.254	0.0953	0.0520	141	-0.83	005.00	0
005	Ash (%)	0354	6.185	0.0100	6.254	0.0953	0.0520	141	-0.73	005.00	0
005	Ash (%)	0848	6.185	0.0300	6.254	0.0953	0.0520	141	-0.73	005.00	0
005	Ash (%)	0897	6.185	0.0100	6.254	0.0953	0.0520	141	-0.73	005.05	0
005	Ash (%)	0066	6.190	0.0200	6.254	0.0953	0.0520	141	-0.68	005.05	0
005	Ash (%)	0512	6.195	0.0690	6.254	0.0953	0.0520	141	-0.63	005.00	0
005	Ash (%)	0108	6.195	0.2100	6.254	0.0953	0.0520	141	-0.62	005.00	0
005	Ash (%)	2089	6.195	0.0700	6.254	0.0953	0.0520	141	-0.62	005.00	0
005	Ash (%)	0726	6.195	0.0100	6.254	0.0953	0.0520	141	-0.62	005.05	0
005	Ash (%)	2109	6.195	0.0500	6.254	0.0953	0.0520	141	-0.62	005.05	0
005	Ash (%)	2097	6.195	0.0083	6.254	0.0953	0.0520	141	-0.62	005.00	0
005	Ash (%)	2076	6.198	0.0293	6.254	0.0953	0.0520	141	-0.60	005.00	0
005	Ash (%)	0910	6.200	0.0000	6.254	0.0953	0.0520	141	-0.57	005.99	0
005	Ash (%)	0171	6.205	0.0500	6.254	0.0953	0.0520	141	-0.52	005.00	0
005	Ash (%)	0650	6.205	0.0100	6.254	0.0953	0.0520	141	-0.52	005.00	0
005	Ash (%)	0027	6.207	0.0260	6.254	0.0953	0.0520	141	-0.50	005.00	0
005	Ash (%)	0546	6.210	0.0200	6.254	0.0953	0.0520	141	-0.47	005.99	0
005	Ash (%)	0199	6.210	0.0400	6.254	0.0953	0.0520	141	-0.47	005.00	0
005	Ash (%)	2268	6.210	0.0400	6.254	0.0953	0.0520	141	-0.47	005.00	0
005	Ash (%)	0263	6.212	0.0040	6.254	0.0953	0.0520	141	-0.44	005.05	0
005	Ash (%)	0035	6.215	0.0100	6.254	0.0953	0.0520	141	-0.41	005.00	0
005	Ash (%)	0651	6.215	0.0000	6.254	0.0953	0.0520	141	-0.41	005.00	0
005	Ash (%)	2193	6.220	0.0000	6.254	0.0953	0.0520	141	-0.36	005.00	0
005	Ash (%)	0042	6.220	0.0600	6.254	0.0953	0.0520	141	-0.36	005.00	0
005	Ash (%)	0164	6.220	0.0200	6.254	0.0953	0.0520	141	-0.36	005.00	0
005	Ash (%)	0529	6.220	0.1400	6.254	0.0953	0.0520	141	-0.36	005.00	0
005	Ash (%)	0589	6.220	0.1200	6.254	0.0953	0.0520	141	-0.36	005.00	0
005	Ash (%)	0148	6.225	0.0300	6.254	0.0953	0.0520	141	-0.31	005.00	0
005	Ash (%)	0520	6.225	0.0100	6.254	0.0953	0.0520	141	-0.31	005.00	0
005	Ash (%)	0964	6.226	0.2430	6.254	0.0953	0.0520	141	-0.30	005.00	0
005	Ash (%)	0510	6.230	0.0400	6.254	0.0953	0.0520	141	-0.26	005.00	0
005	Ash (%)	0511	6.230	0.0600	6.254	0.0953	0.0520	141	-0.26	005.00	0
005	Ash (%)	0638	6.230	0.2200	6.254	0.0953	0.0520	141	-0.26	005.05	0
005	Ash (%)	2012	6.230	0.0600	6.254	0.0953	0.0520	141	-0.26	005.05	0
005	Ash (%)	0202	6.233	0.0100	6.254	0.0953	0.0520	141	-0.22	005.99	0
005	Ash (%)	0622	6.237	0.0514	6.254	0.0953	0.0520	141	-0.18	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0142	6.240	0.0200	6.254	0.0953	0.0520	141	-0.15	005.00	0
005	Ash (%)	0563	6.244	0.1315	6.254	0.0953	0.0520	141	-0.11	005.00	0
005	Ash (%)	0001	6.250	0.0038	6.254	0.0953	0.0520	141	-0.05	005.00	0
005	Ash (%)	0083	6.250	0.0200	6.254	0.0953	0.0520	141	-0.05	005.00	0
005	Ash (%)	0265	6.250	0.1000	6.254	0.0953	0.0520	141	-0.05	005.00	0
005	Ash (%)	0876	6.250	0.1000	6.254	0.0953	0.0520	141	-0.05	005.00	0
005	Ash (%)	0956	6.250	0.1000	6.254	0.0953	0.0520	141	-0.05	005.00	0
005	Ash (%)	0003	6.250	0.0200	6.254	0.0953	0.0520	141	-0.05	005.05	0
005	Ash (%)	0868	6.250	0.0000	6.254	0.0953	0.0520	141	-0.05	005.05	0
005	Ash (%)	2018	6.250	0.0200	6.254	0.0953	0.0520	141	-0.05	005.05	0
005	Ash (%)	0652	6.250	0.1000	6.254	0.0953	0.0520	141	-0.05	005.99	0
005	Ash (%)	0870	6.250	0.0346	6.254	0.0953	0.0520	141	-0.04	005.00	0
005	Ash (%)	0881	6.251	0.0690	6.254	0.0953	0.0520	141	-0.04	005.00	0
005	Ash (%)	0874	6.255	0.0900	6.254	0.0953	0.0520	141	0.01	005.00	0
005	Ash (%)	2188	6.255	0.0300	6.254	0.0953	0.0520	141	0.01	005.00	0
005	Ash (%)	0536	6.255	0.0700	6.254	0.0953	0.0520	141	0.01	005.99	0
005	Ash (%)	0353	6.255	0.0500	6.254	0.0953	0.0520	141	0.01	005.00	0
005	Ash (%)	0298	6.260	0.0200	6.254	0.0953	0.0520	141	0.06	005.00	0
005	Ash (%)	0675	6.260	0.1000	6.254	0.0953	0.0520	141	0.06	005.00	0
005	Ash (%)	0590	6.260	0.0200	6.254	0.0953	0.0520	141	0.06	005.05	0
005	Ash (%)	0849	6.260	0.0200	6.254	0.0953	0.0520	141	0.06	005.05	0
005	Ash (%)	0689	6.270	0.0200	6.254	0.0953	0.0520	141	0.16	005.05	0
005	Ash (%)	2190	6.270	0.1000	6.254	0.0953	0.0520	141	0.16	005.05	0
005	Ash (%)	0505	6.275	0.0700	6.254	0.0953	0.0520	141	0.22	005.00	0
005	Ash (%)	0968	6.277	0.0270	6.254	0.0953	0.0520	141	0.23	005.00	0
005	Ash (%)	0504	6.280	0.0600	6.254	0.0953	0.0520	141	0.27	005.00	0
005	Ash (%)	2058	6.280	0.0200	6.254	0.0953	0.0520	141	0.27	005.00	0
005	Ash (%)	0682	6.280	0.0000	6.254	0.0953	0.0520	141	0.27	005.00	0
005	Ash (%)	2022	6.280	0.0000	6.254	0.0953	0.0520	141	0.27	005.00	0
005	Ash (%)	2248	6.280	0.0000	6.254	0.0953	0.0520	141	0.27	005.00	0
005	Ash (%)	0596	6.280	0.0000	6.254	0.0953	0.0520	141	0.27	005.05	0
005	Ash (%)	0885	6.282	0.0217	6.254	0.0953	0.0520	141	0.29	005.05	0
005	Ash (%)	0695	6.285	0.0100	6.254	0.0953	0.0520	141	0.32	005.05	0
005	Ash (%)	0100	6.290	0.0600	6.254	0.0953	0.0520	141	0.37	005.00	0
005	Ash (%)	0619	6.290	0.0800	6.254	0.0953	0.0520	141	0.37	005.05	0
005	Ash (%)	0098	6.295	0.0500	6.254	0.0953	0.0520	141	0.43	005.00	0
005	Ash (%)	0598	6.295	0.0300	6.254	0.0953	0.0520	141	0.43	005.00	0
005	Ash (%)	2196	6.295	0.0000	6.254	0.0953	0.0520	141	0.43	005.00	0
005	Ash (%)	0017	6.300	0.0000	6.254	0.0953	0.0520	141	0.48	005.00	0
005	Ash (%)	0357	6.300	0.0000	6.254	0.0953	0.0520	141	0.48	005.00	0
005	Ash (%)	0294	6.300	0.0000	6.254	0.0953	0.0520	141	0.48	005.05	0
005	Ash (%)	2246	6.300	0.0400	6.254	0.0953	0.0520	141	0.48	005.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	2009	6.309	0.0207	6.254	0.0953	0.0520	141	0.58	005.05	0
005	Ash (%)	0581	6.315	0.1100	6.254	0.0953	0.0520	141	0.64	005.00	0
005	Ash (%)	2150	6.315	0.0700	6.254	0.0953	0.0520	141	0.64	005.00	0
005	Ash (%)	0610	6.315	0.0100	6.254	0.0953	0.0520	141	0.64	005.05	0
005	Ash (%)	2146	6.315	0.0300	6.254	0.0953	0.0520	141	0.64	005.05	0
005	Ash (%)	2251	6.315	0.0500	6.254	0.0953	0.0520	141	0.64	005.99	0
005	Ash (%)	0889	6.320	0.0059	6.254	0.0953	0.0520	141	0.69	005.05	0
005	Ash (%)	0242	6.320	0.0400	6.254	0.0953	0.0520	141	0.69	005.00	0
005	Ash (%)	0683	6.320	0.2000	6.254	0.0953	0.0520	141	0.69	005.05	0
005	Ash (%)	0676	6.325	0.0100	6.254	0.0953	0.0520	141	0.74	005.99	0
005	Ash (%)	0693	6.333	0.0280	6.254	0.0953	0.0520	141	0.83	005.00	0
005	Ash (%)	0425	6.340	0.0200	6.254	0.0953	0.0520	141	0.90	005.00	0
005	Ash (%)	2053	6.340	0.0200	6.254	0.0953	0.0520	141	0.90	005.00	0
005	Ash (%)	0047	6.350	0.0200	6.254	0.0953	0.0520	141	1.00	005.00	0
005	Ash (%)	0366	6.350	0.1000	6.254	0.0953	0.0520	141	1.00	005.00	0
005	Ash (%)	0884	6.350	0.1000	6.254	0.0953	0.0520	141	1.00	005.99	0
005	Ash (%)	0190	6.355	0.0900	6.254	0.0953	0.0520	141	1.06	005.05	0
005	Ash (%)	2192	6.370	0.0400	6.254	0.0953	0.0520	141	1.21	005.05	0
005	Ash (%)	2073	6.370	0.0000	6.254	0.0953	0.0520	141	1.21	005.00	0
005	Ash (%)	2062	6.378	0.0005	6.254	0.0953	0.0520	141	1.29	005.05	0
005	Ash (%)	0720	6.380	0.0200	6.254	0.0953	0.0520	141	1.32	005.05	0
005	Ash (%)	0948	6.380	0.0400	6.254	0.0953	0.0520	141	1.32	005.00	0
005	Ash (%)	0941	6.380	0.0400	6.254	0.0953	0.0520	141	1.32	005.05	0
005	Ash (%)	0139	6.385	0.0300	6.254	0.0953	0.0520	141	1.37	005.00	0
005	Ash (%)	0644	6.385	0.0100	6.254	0.0953	0.0520	141	1.37	005.02	0
005	Ash (%)	0889	6.393	0.2292	6.254	0.0953	0.0520	141	1.45	005.11	0
005	Ash (%)	0685	6.410	0.1000	6.254	0.0953	0.0520	141	1.63	005.05	0
005	Ash (%)	2198	6.410	0.0000	6.254	0.0953	0.0520	141	1.63	005.05	0
005	Ash (%)	2262	6.410	0.0000	6.254	0.0953	0.0520	141	1.63	005.05	0
005	Ash (%)	2129	6.430	0.0800	6.254	0.0953	0.0520	141	1.84	005.05	0
005	Ash (%)	0939	6.445	0.0100	6.254	0.0953	0.0520	141	2.00	005.05	0
005	Ash (%)	0674	6.460	0.0000	6.254	0.0953	0.0520	141	2.16	005.00	0
005	Ash (%)	0852	6.500	0.0000	6.254	0.0953	0.0520	141	2.58	005.00	0
005	Ash (%)	0904	6.575	0.2500	6.254	0.0953	0.0520	141	3.37	005.05	0
005	Ash (%)	0226	6.885	0.2500	6.254	0.0953	0.0520	141	6.62	005.00	0
005	Ash (%)	2259	6.907	0.1730	6.254	0.0953	0.0520	141	6.84	005.00	0
005	Ash (%)	0852	7.445	0.0900	6.254	0.0953	0.0520	141	12.50	005.11	0
005	Ash (%)	2128	9.003	0.1704	6.254	0.0953	0.0520	141	28.85	005.11	0
005	Ash (%)	0045	5.930	0.4600	6.254	0.0953	0.0520	141	-3.40	005.00	1
005	Ash (%)	2081	6.560	0.9000	6.254	0.0953	0.0520	141	3.21	005.99	1
005	Ash (%)	0034	9.279	1.040	6.254	0.0953	0.0520	141	31.74	005.00	2
005	Ash (%)	0062	9.386	0.3132	6.254	0.0953	0.0520	141	32.86	005.00	2

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
006	Total Sugars (%)	0720	4.230	0.0200	6.330	1.627	0.3165	4		006.99	0
006	Total Sugars (%)	0910	6.390	0.2400	6.330	1.627	0.3165	4		006.00	0
006	Total Sugars (%)	0956	6.500	0.4000	6.330	1.627	0.3165	4		006.99	0
006	Total Sugars (%)	0407	8.200	0.6060	6.330	1.627	0.3165	4		006.00	0
008	Fiber, Acid Detergent (%)	0956	8.300	1.400	10.16	0.9709	0.3842	47	-1.92	008.08	0
008	Fiber, Acid Detergent (%)	0918	8.825	0.6363	10.16	0.9709	0.3842	47	-1.38	008.08	0
008	Fiber, Acid Detergent (%)	0590	8.845	0.0700	10.16	0.9709	0.3842	47	-1.36	008.08	0
008	Fiber, Acid Detergent (%)	0049	8.990	0.7800	10.16	0.9709	0.3842	47	-1.21	008.08	0
008	Fiber, Acid Detergent (%)	0619	9.055	0.0500	10.16	0.9709	0.3842	47	-1.14	008.08	0
008	Fiber, Acid Detergent (%)	2012	9.060	0.6000	10.16	0.9709	0.3842	47	-1.14	008.08	0
008	Fiber, Acid Detergent (%)	2190	9.225	1.250	10.16	0.9709	0.3842	47	-0.97	008.08	0
008	Fiber, Acid Detergent (%)	0910	9.270	0.5400	10.16	0.9709	0.3842	47	-0.92	008.08	0
008	Fiber, Acid Detergent (%)	0164	9.300	0.2000	10.16	0.9709	0.3842	47	-0.89	008.08	0
008	Fiber, Acid Detergent (%)	0148	9.370	0.3400	10.16	0.9709	0.3842	47	-0.82	008.08	0
008	Fiber, Acid Detergent (%)	0504	9.420	0.5600	10.16	0.9709	0.3842	47	-0.77	008.02	0
008	Fiber, Acid Detergent (%)	0098	9.485	0.2500	10.16	0.9709	0.3842	47	-0.70	008.02	0
008	Fiber, Acid Detergent (%)	0939	9.500	0.7000	10.16	0.9709	0.3842	47	-0.68	008.08	0
008	Fiber, Acid Detergent (%)	0354	9.520	0.0000	10.16	0.9709	0.3842	47	-0.66	008.08	0
008	Fiber, Acid Detergent (%)	0263	9.540	0.3828	10.16	0.9709	0.3842	47	-0.64	008.08	0
008	Fiber, Acid Detergent (%)	2196	9.660	0.0000	10.16	0.9709	0.3842	47	-0.52	008.08	0
008	Fiber, Acid Detergent (%)	0720	9.675	0.1300	10.16	0.9709	0.3842	47	-0.50	008.99	0
008	Fiber, Acid Detergent (%)	0353	9.825	0.2500	10.16	0.9709	0.3842	47	-0.35	008.02	0
008	Fiber, Acid Detergent (%)	0407	9.860	0.9789	10.16	0.9709	0.3842	47	-0.31	008.08	0
008	Fiber, Acid Detergent (%)	0358	9.885	0.5900	10.16	0.9709	0.3842	47	-0.29	008.08	0
008	Fiber, Acid Detergent (%)	0294	9.900	0.0000	10.16	0.9709	0.3842	47	-0.27	008.08	0
008	Fiber, Acid Detergent (%)	0870	10.09	0.2489	10.16	0.9709	0.3842	47	-0.08	008.08	0
008	Fiber, Acid Detergent (%)	0042	10.10	0.1500	10.16	0.9709	0.3842	47	-0.07	008.08	0
008	Fiber, Acid Detergent (%)	0357	10.10	0.4000	10.16	0.9709	0.3842	47	-0.06	008.08	0
008	Fiber, Acid Detergent (%)	0510	10.20	0.2000	10.16	0.9709	0.3842	47	0.04	008.08	0
008	Fiber, Acid Detergent (%)	0675	10.32	0.2100	10.16	0.9709	0.3842	47	0.16	008.08	0
008	Fiber, Acid Detergent (%)	0001	10.32	0.1000	10.16	0.9709	0.3842	47	0.16	008.08	0
008	Fiber, Acid Detergent (%)	0265	10.35	0.5000	10.16	0.9709	0.3842	47	0.19	008.05	0
008	Fiber, Acid Detergent (%)	0278	10.35	0.1000	10.16	0.9709	0.3842	47	0.19	008.08	0
008	Fiber, Acid Detergent (%)	0148	10.42	0.3600	10.16	0.9709	0.3842	47	0.26	008.02	0
008	Fiber, Acid Detergent (%)	2193	10.47	0.2700	10.16	0.9709	0.3842	47	0.31	008.08	0
008	Fiber, Acid Detergent (%)	0066	10.50	0.4500	10.16	0.9709	0.3842	47	0.34	008.08	0
008	Fiber, Acid Detergent (%)	0675	10.53	0.1200	10.16	0.9709	0.3842	47	0.38	008.02	0
008	Fiber, Acid Detergent (%)	0968	10.55	0.2490	10.16	0.9709	0.3842	47	0.39	008.08	0
008	Fiber, Acid Detergent (%)	2192	10.66	0.0500	10.16	0.9709	0.3842	47	0.51	008.02	0
008	Fiber, Acid Detergent (%)	0581	10.70	0.2200	10.16	0.9709	0.3842	47	0.55	008.08	0
008	Fiber, Acid Detergent (%)	0083	10.92	0.3400	10.16	0.9709	0.3842	47	0.78	008.08	0
008	Fiber, Acid Detergent (%)	0726	10.97	0.0500	10.16	0.9709	0.3842	47	0.83	008.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
008	Fiber, Acid Detergent (%)	0948	11.08	0.2800	10.16	0.9709	0.3842	47	0.94	008.02	0
008	Fiber, Acid Detergent (%)	0226	11.29	0.0300	10.16	0.9709	0.3842	47	1.16	008.02	0
008	Fiber, Acid Detergent (%)	0536	11.38	0.3100	10.16	0.9709	0.3842	47	1.25	008.08	0
008	Fiber, Acid Detergent (%)	0190	11.45	0.5000	10.16	0.9709	0.3842	47	1.33	008.08	0
008	Fiber, Acid Detergent (%)	0646	11.57	0.4300	10.16	0.9709	0.3842	47	1.44	008.08	0
008	Fiber, Acid Detergent (%)	0693	11.95	0.4400	10.16	0.9709	0.3842	47	1.84	008.08	0
008	Fiber, Acid Detergent (%)	0045	12.40	0.2000	10.16	0.9709	0.3842	47	2.30	008.08	0
008	Fiber, Acid Detergent (%)	0948	12.48	1.730	10.16	0.9709	0.3842	47	2.38	008.08	0
008	Fiber, Acid Detergent (%)	0037	13.47	0.4100	10.16	0.9709	0.3842	47	3.40	008.08	0
008	Fiber, Acid Detergent (%)	0889	6.249	12.50	10.16	0.9709	0.3842	47		008.08	4
009	Fiber, Neutral Detergent (%)	0939	19.96	0.5700	22.60	2.178	0.5478	48	-1.22	009.09	0
009	Fiber, Neutral Detergent (%)	2012	20.10	0.0700	22.60	2.178	0.5478	48	-1.15	009.09	0
009	Fiber, Neutral Detergent (%)	0263	20.10	0.2944	22.60	2.178	0.5478	48	-1.15	009.09	0
009	Fiber, Neutral Detergent (%)	0278	20.20	0.6000	22.60	2.178	0.5478	48	-1.10	009.09	0
009	Fiber, Neutral Detergent (%)	0619	20.20	0.2000	22.60	2.178	0.5478	48	-1.10	009.09	0
009	Fiber, Neutral Detergent (%)	2190	20.22	0.8000	22.60	2.178	0.5478	48	-1.09	009.09	0
009	Fiber, Neutral Detergent (%)	0357	20.35	1.500	22.60	2.178	0.5478	48	-1.03	009.09	0
009	Fiber, Neutral Detergent (%)	0598	20.39	0.3600	22.60	2.178	0.5478	48	-1.02	009.09	0
009	Fiber, Neutral Detergent (%)	0956	20.55	2.500	22.60	2.178	0.5478	48	-0.94	009.09	0
009	Fiber, Neutral Detergent (%)	0918	21.03	0.5565	22.60	2.178	0.5478	48	-0.72	009.09	0
009	Fiber, Neutral Detergent (%)	0504	21.11	0.0100	22.60	2.178	0.5478	48	-0.69	009.04	0
009	Fiber, Neutral Detergent (%)	0358	21.13	0.0100	22.60	2.178	0.5478	48	-0.68	009.09	0
009	Fiber, Neutral Detergent (%)	0581	21.20	0.0100	22.60	2.178	0.5478	48	-0.65	009.09	0
009	Fiber, Neutral Detergent (%)	0098	21.20	3.800	22.60	2.178	0.5478	48	-0.64	009.07	0
009	Fiber, Neutral Detergent (%)	0049	21.35	1.990	22.60	2.178	0.5478	48	-0.58	009.09	0
009	Fiber, Neutral Detergent (%)	0164	21.35	0.1000	22.60	2.178	0.5478	48	-0.57	009.09	0
009	Fiber, Neutral Detergent (%)	0353	21.37	0.2500	22.60	2.178	0.5478	48	-0.57	009.07	0
009	Fiber, Neutral Detergent (%)	2193	21.47	0.1400	22.60	2.178	0.5478	48	-0.52	009.09	0
009	Fiber, Neutral Detergent (%)	0354	21.48	0.0000	22.60	2.178	0.5478	48	-0.51	009.09	0
009	Fiber, Neutral Detergent (%)	0870	21.60	0.9750	22.60	2.178	0.5478	48	-0.46	009.09	0
009	Fiber, Neutral Detergent (%)	0910	22.00	0.2000	22.60	2.178	0.5478	48	-0.28	009.07	0
009	Fiber, Neutral Detergent (%)	0407	22.01	0.2923	22.60	2.178	0.5478	48	-0.27	009.09	0
009	Fiber, Neutral Detergent (%)	0001	22.04	0.5700	22.60	2.178	0.5478	48	-0.26	009.09	0
009	Fiber, Neutral Detergent (%)	0510	22.20	0.0000	22.60	2.178	0.5478	48	-0.18	009.09	0
009	Fiber, Neutral Detergent (%)	0045	22.30	0.0000	22.60	2.178	0.5478	48	-0.14	009.09	0
009	Fiber, Neutral Detergent (%)	0889	22.35	0.1329	22.60	2.178	0.5478	48	-0.12	009.99	0
009	Fiber, Neutral Detergent (%)	0294	22.45	0.1000	22.60	2.178	0.5478	48	-0.07	009.09	0
009	Fiber, Neutral Detergent (%)	0226	22.48	0.6500	22.60	2.178	0.5478	48	-0.06	009.07	0
009	Fiber, Neutral Detergent (%)	2196	22.61	0.0000	22.60	2.178	0.5478	48	0.00	009.09	0
009	Fiber, Neutral Detergent (%)	0646	22.74	0.9700	22.60	2.178	0.5478	48	0.06	009.09	0
009	Fiber, Neutral Detergent (%)	0948	23.15	1.210	22.60	2.178	0.5478	48	0.25	009.07	0
009	Fiber, Neutral Detergent (%)	0148	23.40	1.400	22.60	2.178	0.5478	48	0.37	009.09	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
009	Fiber, Neutral Detergent (%)	0638	23.50	0.2000	22.60	2.178	0.5478	48	0.41	009.09	0
009	Fiber, Neutral Detergent (%)	0948	23.63	0.4000	22.60	2.178	0.5478	48	0.47	009.09	0
009	Fiber, Neutral Detergent (%)	0083	23.72	1.060	22.60	2.178	0.5478	48	0.51	009.09	0
009	Fiber, Neutral Detergent (%)	2192	23.78	0.1700	22.60	2.178	0.5478	48	0.54	009.07	0
009	Fiber, Neutral Detergent (%)	0693	23.91	0.8360	22.60	2.178	0.5478	48	0.60	009.09	0
009	Fiber, Neutral Detergent (%)	0066	24.04	0.4000	22.60	2.178	0.5478	48	0.66	009.09	0
009	Fiber, Neutral Detergent (%)	0590	24.84	0.6100	22.60	2.178	0.5478	48	1.03	009.09	0
009	Fiber, Neutral Detergent (%)	0536	24.98	0.8700	22.60	2.178	0.5478	48	1.09	009.09	0
009	Fiber, Neutral Detergent (%)	0037	25.53	0.1300	22.60	2.178	0.5478	48	1.34	009.09	0
009	Fiber, Neutral Detergent (%)	0265	25.70	0.4000	22.60	2.178	0.5478	48	1.42	009.09	0
009	Fiber, Neutral Detergent (%)	0968	25.89	0.1680	22.60	2.178	0.5478	48	1.51	009.09	0
009	Fiber, Neutral Detergent (%)	0675	26.23	0.4300	22.60	2.178	0.5478	48	1.66	009.09	0
009	Fiber, Neutral Detergent (%)	0720	26.28	0.0900	22.60	2.178	0.5478	48	1.69	009.99	0
009	Fiber, Neutral Detergent (%)	0675	26.72	0.1300	22.60	2.178	0.5478	48	1.89	009.07	0
009	Fiber, Neutral Detergent (%)	0190	30.35	0.1100	22.60	2.178	0.5478	48	3.56	009.09	0
009	Fiber, Neutral Detergent (%)	0726	31.55	0.0300	22.60	2.178	0.5478	48	4.11	009.09	0
009	Fiber, Neutral Detergent (%)	0889	10.47	20.95	22.60	2.178	0.5478	48		009.09	4
010	Moisture (%)	0164	6.495	0.0100	7.860	0.8026	0.1058	22	-1.70	010.03	0
010	Moisture (%)	0546	6.815	0.0500	7.860	0.8026	0.1058	22	-1.30	010.99	0
010	Moisture (%)	0726	7.220	0.0200	7.860	0.8026	0.1058	22	-0.80	010.99	0
010	Moisture (%)	0939	7.360	0.2000	7.860	0.8026	0.1058	22	-0.62	010.99	0
010	Moisture (%)	0843	7.440	0.0400	7.860	0.8026	0.1058	22	-0.52	010.03	0
010	Moisture (%)	0884	7.450	0.3000	7.860	0.8026	0.1058	22	-0.51	010.99	0
010	Moisture (%)	2089	7.460	0.0000	7.860	0.8026	0.1058	22	-0.50	010.99	0
010	Moisture (%)	0948	7.490	0.1000	7.860	0.8026	0.1058	22	-0.46	010.99	0
010	Moisture (%)	0652	7.500	0.4000	7.860	0.8026	0.1058	22	-0.45	010.99	0
010	Moisture (%)	0553	7.525	0.0300	7.860	0.8026	0.1058	22	-0.42	010.11	0
010	Moisture (%)	2043	7.640	0.1200	7.860	0.8026	0.1058	22	-0.27	010.99	0
010	Moisture (%)	2251	7.675	0.2900	7.860	0.8026	0.1058	22	-0.23	010.99	0
010	Moisture (%)	2150	7.735	0.1900	7.860	0.8026	0.1058	22	-0.16	010.99	0
010	Moisture (%)	2178	7.835	0.3500	7.860	0.8026	0.1058	22	-0.03	010.99	0
010	Moisture (%)	2076	7.960	0.0200	7.860	0.8026	0.1058	22	0.12	010.99	0
010	Moisture (%)	0190	8.020	0.0200	7.860	0.8026	0.1058	22	0.20	010.99	0
010	Moisture (%)	0889	8.355	0.0000	7.860	0.8026	0.1058	22	0.62	010.11	0
010	Moisture (%)	2248	8.820	0.0200	7.860	0.8026	0.1058	22	1.20	010.99	0
010	Moisture (%)	0720	8.855	0.0100	7.860	0.8026	0.1058	22	1.24	010.11	0
010	Moisture (%)	2128	8.976	0.1277	7.860	0.8026	0.1058	22	1.39	010.11	0
010	Moisture (%)	0405	9.375	0.0100	7.860	0.8026	0.1058	22	1.89	010.99	0
010	Moisture (%)	0852	10.04	0.0200	7.860	0.8026	0.1058	22	2.72	010.11	0
010	Moisture (%)	2129	7.400	0.6000	7.860	0.8026	0.1058	22	-0.57	010.99	1
011	Loss on Drying (%)	2256	7.355	0.1900	8.307	0.3513	0.0856	69	-2.71	011.01	0
011	Loss on Drying (%)	0685	7.590	0.2200	8.307	0.3513	0.0856	69	-2.04	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
011	Loss on Drying (%)	0650	7.610	0.1200	8.307	0.3513	0.0856	69	-1.98	011.01	0
011	Loss on Drying (%)	0536	7.615	0.0100	8.307	0.3513	0.0856	69	-1.97	011.01	0
011	Loss on Drying (%)	0874	7.625	0.0700	8.307	0.3513	0.0856	69	-1.94	011.01	0
011	Loss on Drying (%)	0062	7.717	0.0860	8.307	0.3513	0.0856	69	-1.68	011.01	0
011	Loss on Drying (%)	0859	7.750	0.0020	8.307	0.3513	0.0856	69	-1.59	011.01	0
011	Loss on Drying (%)	2168	7.765	0.2100	8.307	0.3513	0.0856	69	-1.54	011.01	0
011	Loss on Drying (%)	0563	7.829	0.0989	8.307	0.3513	0.0856	69	-1.36	011.01	0
011	Loss on Drying (%)	0539	7.880	0.0600	8.307	0.3513	0.0856	69	-1.22	011.01	0
011	Loss on Drying (%)	0942	7.890	0.3600	8.307	0.3513	0.0856	69	-1.19	011.02	0
011	Loss on Drying (%)	0968	7.894	0.0330	8.307	0.3513	0.0856	69	-1.18	011.01	0
011	Loss on Drying (%)	0953	7.940	0.0600	8.307	0.3513	0.0856	69	-1.04	011.01	0
011	Loss on Drying (%)	0622	8.067	0.1380	8.307	0.3513	0.0856	69	-0.68	011.01	0
011	Loss on Drying (%)	0646	8.070	0.2000	8.307	0.3513	0.0856	69	-0.67	011.01	0
011	Loss on Drying (%)	2198	8.075	0.0100	8.307	0.3513	0.0856	69	-0.66	011.01	0
011	Loss on Drying (%)	0510	8.100	0.0000	8.307	0.3513	0.0856	69	-0.59	011.01	0
011	Loss on Drying (%)	0425	8.125	0.0500	8.307	0.3513	0.0856	69	-0.52	011.01	0
011	Loss on Drying (%)	2033	8.145	0.0100	8.307	0.3513	0.0856	69	-0.46	011.01	0
011	Loss on Drying (%)	0598	8.165	0.0100	8.307	0.3513	0.0856	69	-0.40	011.01	0
011	Loss on Drying (%)	0407	8.167	0.0181	8.307	0.3513	0.0856	69	-0.40	011.01	0
011	Loss on Drying (%)	2073	8.170	0.0200	8.307	0.3513	0.0856	69	-0.39	011.01	0
011	Loss on Drying (%)	0098	8.200	0.0800	8.307	0.3513	0.0856	69	-0.30	011.01	0
011	Loss on Drying (%)	0857	8.200	0.0000	8.307	0.3513	0.0856	69	-0.30	011.99	0
011	Loss on Drying (%)	0881	8.207	0.1600	8.307	0.3513	0.0856	69	-0.28	011.01	0
011	Loss on Drying (%)	2268	8.240	0.1000	8.307	0.3513	0.0856	69	-0.19	011.01	0
011	Loss on Drying (%)	0164	8.245	0.1100	8.307	0.3513	0.0856	69	-0.18	011.01	0
011	Loss on Drying (%)	0047	8.270	0.3400	8.307	0.3513	0.0856	69	-0.11	011.01	0
011	Loss on Drying (%)	0354	8.290	0.0200	8.307	0.3513	0.0856	69	-0.05	011.01	0
011	Loss on Drying (%)	2188	8.290	0.1200	8.307	0.3513	0.0856	69	-0.05	011.01	0
011	Loss on Drying (%)	0265	8.300	0.0000	8.307	0.3513	0.0856	69	-0.02	011.01	0
011	Loss on Drying (%)	0294	8.300	0.0000	8.307	0.3513	0.0856	69	-0.02	011.01	0
011	Loss on Drying (%)	0263	8.325	0.0100	8.307	0.3513	0.0856	69	0.05	011.01	0
011	Loss on Drying (%)	0108	8.330	0.0000	8.307	0.3513	0.0856	69	0.07	011.01	0
011	Loss on Drying (%)	0242	8.340	0.0800	8.307	0.3513	0.0856	69	0.09	011.01	0
011	Loss on Drying (%)	0171	8.345	0.0700	8.307	0.3513	0.0856	69	0.11	011.01	0
011	Loss on Drying (%)	0226	8.350	0.0000	8.307	0.3513	0.0856	69	0.12	011.01	0
011	Loss on Drying (%)	2184	8.350	0.0400	8.307	0.3513	0.0856	69	0.12	011.01	0
011	Loss on Drying (%)	2006	8.370	0.0600	8.307	0.3513	0.0856	69	0.18	011.01	0
011	Loss on Drying (%)	0880	8.375	0.3300	8.307	0.3513	0.0856	69	0.19	011.01	0
011	Loss on Drying (%)	0229	8.394	0.0260	8.307	0.3513	0.0856	69	0.25	011.01	0
011	Loss on Drying (%)	0870	8.394	0.0674	8.307	0.3513	0.0856	69	0.25	011.01	0
011	Loss on Drying (%)	0144	8.405	0.0100	8.307	0.3513	0.0856	69	0.28	011.01	0
011	Loss on Drying (%)	0298	8.420	0.0200	8.307	0.3513	0.0856	69	0.32	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
011	Loss on Drying (%)	0511	8.420	0.1400	8.307	0.3513	0.0856	69	0.32	011.01	0
011	Loss on Drying (%)	0574	8.430	0.0400	8.307	0.3513	0.0856	69	0.35	011.01	0
011	Loss on Drying (%)	2196	8.430	0.0000	8.307	0.3513	0.0856	69	0.35	011.01	0
011	Loss on Drying (%)	2022	8.440	0.0800	8.307	0.3513	0.0856	69	0.38	011.01	0
011	Loss on Drying (%)	2193	8.450	0.0800	8.307	0.3513	0.0856	69	0.41	011.01	0
011	Loss on Drying (%)	0843	8.460	0.0200	8.307	0.3513	0.0856	69	0.44	011.01	0
011	Loss on Drying (%)	0100	8.480	0.0800	8.307	0.3513	0.0856	69	0.49	011.01	0
011	Loss on Drying (%)	0175	8.500	0.2000	8.307	0.3513	0.0856	69	0.55	011.01	0
011	Loss on Drying (%)	0674	8.520	0.0000	8.307	0.3513	0.0856	69	0.61	011.01	0
011	Loss on Drying (%)	0682	8.540	0.0000	8.307	0.3513	0.0856	69	0.66	011.01	0
011	Loss on Drying (%)	0541	8.565	0.4300	8.307	0.3513	0.0856	69	0.73	011.01	0
011	Loss on Drying (%)	0004	8.595	0.0300	8.307	0.3513	0.0856	69	0.82	011.01	0
011	Loss on Drying (%)	2097	8.605	0.0100	8.307	0.3513	0.0856	69	0.85	011.01	0
011	Loss on Drying (%)	0675	8.615	0.0900	8.307	0.3513	0.0856	69	0.88	011.01	0
011	Loss on Drying (%)	2031	8.630	0.0296	8.307	0.3513	0.0856	69	0.92	011.01	0
011	Loss on Drying (%)	0643	8.650	0.3000	8.307	0.3513	0.0856	69	0.98	011.01	0
011	Loss on Drying (%)	0589	8.680	0.1200	8.307	0.3513	0.0856	69	1.06	011.01	0
011	Loss on Drying (%)	0202	8.756	0.1670	8.307	0.3513	0.0856	69	1.28	011.01	0
011	Loss on Drying (%)	0066	8.765	0.1300	8.307	0.3513	0.0856	69	1.30	011.01	0
011	Loss on Drying (%)	2058	8.780	0.0600	8.307	0.3513	0.0856	69	1.35	011.01	0
011	Loss on Drying (%)	0520	8.900	0.0000	8.307	0.3513	0.0856	69	1.69	011.01	0
011	Loss on Drying (%)	0651	8.924	0.1330	8.307	0.3513	0.0856	69	1.76	011.01	0
011	Loss on Drying (%)	0618	8.975	0.0100	8.307	0.3513	0.0856	69	1.90	011.01	0
011	Loss on Drying (%)	0358	9.000	0.0600	8.307	0.3513	0.0856	69	1.97	011.01	0
011	Loss on Drying (%)	0852	9.220	0.0800	8.307	0.3513	0.0856	69	2.60	011.99	0
011	Loss on Drying (%)	0553	8.095	0.6500	8.307	0.3513	0.0856	69	-0.60	011.01	1
012	Starch (%)	0918	6.417	0.4800	11.85	1.778	0.2245	36	-3.05	012.01	0
012	Starch (%)	0164	8.500	0.6000	11.85	1.778	0.2245	36	-1.88	012.01	0
012	Starch (%)	0278	9.500	0.2000	11.85	1.778	0.2245	36	-1.32	012.04	0
012	Starch (%)	0510	9.500	0.4000	11.85	1.778	0.2245	36	-1.32	012.04	0
012	Starch (%)	0870	9.752	0.2115	11.85	1.778	0.2245	36	-1.18	012.01	0
012	Starch (%)	0676	10.03	0.0400	11.85	1.778	0.2245	36	-1.02	012.01	0
012	Starch (%)	0956	10.15	0.1000	11.85	1.778	0.2245	36	-0.95	012.01	0
012	Starch (%)	0148	10.24	0.4000	11.85	1.778	0.2245	36	-0.90	012.01	0
012	Starch (%)	0407	10.44	0.5638	11.85	1.778	0.2245	36	-0.79	012.03	0
012	Starch (%)	0353	10.50	0.0700	11.85	1.778	0.2245	36	-0.76	012.04	0
012	Starch (%)	0045	10.50	0.8000	11.85	1.778	0.2245	36	-0.76	012.01	0
012	Starch (%)	0948	10.83	0.3900	11.85	1.778	0.2245	36	-0.58	012.04	0
012	Starch (%)	0098	10.90	0.0000	11.85	1.778	0.2245	36	-0.53	012.03	0
012	Starch (%)	0066	11.23	0.1400	11.85	1.778	0.2245	36	-0.35	012.03	0
012	Starch (%)	0164	11.80	0.2000	11.85	1.778	0.2245	36	-0.03	012.00	0
012	Starch (%)	2188	11.89	0.3800	11.85	1.778	0.2245	36	0.02	012.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
012	Starch (%)	0610	11.90	0.0000	11.85	1.778	0.2245	36	0.03	012.00	0
012	Starch (%)	0889	11.94	0.4995	11.85	1.778	0.2245	36	0.05	012.11	0
012	Starch (%)	0889	12.03	0.0163	11.85	1.778	0.2245	36	0.10	012.03	0
012	Starch (%)	2009	12.21	0.0255	11.85	1.778	0.2245	36	0.20	012.00	0
012	Starch (%)	2006	12.25	0.0100	11.85	1.778	0.2245	36	0.22	012.00	0
012	Starch (%)	0619	12.30	0.0000	11.85	1.778	0.2245	36	0.25	012.00	0
012	Starch (%)	2196	12.72	0.0000	11.85	1.778	0.2245	36	0.49	012.00	0
012	Starch (%)	0689	12.90	0.2000	11.85	1.778	0.2245	36	0.59	012.00	0
012	Starch (%)	2193	13.12	0.1200	11.85	1.778	0.2245	36	0.71	012.00	0
012	Starch (%)	0354	13.14	0.2400	11.85	1.778	0.2245	36	0.73	012.00	0
012	Starch (%)	0939	13.27	0.1100	11.85	1.778	0.2245	36	0.80	012.00	0
012	Starch (%)	2128	13.30	0.3350	11.85	1.778	0.2245	36	0.81	012.11	0
012	Starch (%)	0910	13.30	0.4000	11.85	1.778	0.2245	36	0.82	012.00	0
012	Starch (%)	2192	13.43	0.0900	11.85	1.778	0.2245	36	0.89	012.00	0
012	Starch (%)	0265	13.45	0.3000	11.85	1.778	0.2245	36	0.90	012.01	0
012	Starch (%)	0941	13.48	0.2200	11.85	1.778	0.2245	36	0.92	012.00	0
012	Starch (%)	2089	13.62	0.3500	11.85	1.778	0.2245	36	0.99	012.01	0
012	Starch (%)	0683	13.73	0.1500	11.85	1.778	0.2245	36	1.06	012.00	0
012	Starch (%)	0720	14.36	0.0400	11.85	1.778	0.2245	36	1.41	012.11	0
012	Starch (%)	0638	15.20	0.0000	11.85	1.778	0.2245	36	1.88	012.00	0
012	Starch (%)	0693	12.83	1.501	11.85	1.778	0.2245	36	0.55	012.01	1
013	Fat, Acid Pretreat (%)	0504	1.635	0.0500	6.664	0.8099	0.1818	57	-6.21	013.00	0
013	Fat, Acid Pretreat (%)	0036	3.760	0.0906	6.664	0.8099	0.1818	57	-3.59	013.10	0
013	Fat, Acid Pretreat (%)	0720	4.145	0.0300	6.664	0.8099	0.1818	57	-3.11	013.10	0
013	Fat, Acid Pretreat (%)	0353	4.350	0.4000	6.664	0.8099	0.1818	57	-2.86	013.10	0
013	Fat, Acid Pretreat (%)	0618	4.668	0.1127	6.664	0.8099	0.1818	57	-2.46	013.08	0
013	Fat, Acid Pretreat (%)	0083	5.165	0.0100	6.664	0.8099	0.1818	57	-1.85	013.00	0
013	Fat, Acid Pretreat (%)	0689	5.700	0.0000	6.664	0.8099	0.1818	57	-1.19	013.00	0
013	Fat, Acid Pretreat (%)	0939	5.775	0.1300	6.664	0.8099	0.1818	57	-1.10	013.10	0
013	Fat, Acid Pretreat (%)	0618	5.817	0.3156	6.664	0.8099	0.1818	57	-1.05	013.00	0
013	Fat, Acid Pretreat (%)	0505	5.825	0.2300	6.664	0.8099	0.1818	57	-1.04	013.13	0
013	Fat, Acid Pretreat (%)	0407	5.827	0.0455	6.664	0.8099	0.1818	57	-1.03	013.02	0
013	Fat, Acid Pretreat (%)	2073	5.880	0.0200	6.664	0.8099	0.1818	57	-0.97	013.00	0
013	Fat, Acid Pretreat (%)	0853	5.975	0.6300	6.664	0.8099	0.1818	57	-0.85	013.02	0
013	Fat, Acid Pretreat (%)	0638	6.050	0.1000	6.664	0.8099	0.1818	57	-0.76	013.10	0
013	Fat, Acid Pretreat (%)	0650	6.300	0.6000	6.664	0.8099	0.1818	57	-0.45	013.02	0
013	Fat, Acid Pretreat (%)	0884	6.300	0.2000	6.664	0.8099	0.1818	57	-0.45	013.00	0
013	Fat, Acid Pretreat (%)	0242	6.320	0.1600	6.664	0.8099	0.1818	57	-0.42	013.00	0
013	Fat, Acid Pretreat (%)	0229	6.325	0.1100	6.664	0.8099	0.1818	57	-0.42	013.00	0
013	Fat, Acid Pretreat (%)	2196	6.350	0.0000	6.664	0.8099	0.1818	57	-0.39	013.10	0
013	Fat, Acid Pretreat (%)	0610	6.380	0.3400	6.664	0.8099	0.1818	57	-0.35	013.10	0
013	Fat, Acid Pretreat (%)	0581	6.430	0.2200	6.664	0.8099	0.1818	57	-0.29	013.13	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
013	Fat, Acid Pretreat (%)	0171	6.460	0.0600	6.664	0.8099	0.1818	57	-0.25	013.02	0
013	Fat, Acid Pretreat (%)	0910	6.530	0.2800	6.664	0.8099	0.1818	57	-0.17	013.00	0
013	Fat, Acid Pretreat (%)	0948	6.585	0.0700	6.664	0.8099	0.1818	57	-0.10	013.00	0
013	Fat, Acid Pretreat (%)	0874	6.610	0.0400	6.664	0.8099	0.1818	57	-0.07	013.02	0
013	Fat, Acid Pretreat (%)	0870	6.627	0.3907	6.664	0.8099	0.1818	57	-0.05	013.02	0
013	Fat, Acid Pretreat (%)	0652	6.650	0.1000	6.664	0.8099	0.1818	57	-0.02	013.00	0
013	Fat, Acid Pretreat (%)	0354	6.670	0.2000	6.664	0.8099	0.1818	57	0.01	013.02	0
013	Fat, Acid Pretreat (%)	0098	6.710	0.4200	6.664	0.8099	0.1818	57	0.06	013.02	0
013	Fat, Acid Pretreat (%)	2259	6.714	0.3610	6.664	0.8099	0.1818	57	0.06	013.02	0
013	Fat, Acid Pretreat (%)	2076	6.744	0.4884	6.664	0.8099	0.1818	57	0.10	013.00	0
013	Fat, Acid Pretreat (%)	0148	6.755	0.3300	6.664	0.8099	0.1818	57	0.11	013.02	0
013	Fat, Acid Pretreat (%)	0843	6.805	0.2500	6.664	0.8099	0.1818	57	0.17	013.13	0
013	Fat, Acid Pretreat (%)	0004	6.855	0.6700	6.664	0.8099	0.1818	57	0.24	013.02	0
013	Fat, Acid Pretreat (%)	0541	6.900	0.0400	6.664	0.8099	0.1818	57	0.29	013.00	0
013	Fat, Acid Pretreat (%)	0904	6.905	0.0300	6.664	0.8099	0.1818	57	0.30	013.13	0
013	Fat, Acid Pretreat (%)	0100	6.935	0.2500	6.664	0.8099	0.1818	57	0.33	013.02	0
013	Fat, Acid Pretreat (%)	0035	6.935	0.0700	6.664	0.8099	0.1818	57	0.33	013.00	0
013	Fat, Acid Pretreat (%)	0074	6.935	0.0100	6.664	0.8099	0.1818	57	0.33	013.02	0
013	Fat, Acid Pretreat (%)	0017	7.100	0.0000	6.664	0.8099	0.1818	57	0.54	013.02	0
013	Fat, Acid Pretreat (%)	0921	7.105	0.0100	6.664	0.8099	0.1818	57	0.54	013.02	0
013	Fat, Acid Pretreat (%)	0015	7.155	0.1300	6.664	0.8099	0.1818	57	0.61	013.13	0
013	Fat, Acid Pretreat (%)	0139	7.245	0.0700	6.664	0.8099	0.1818	57	0.72	013.00	0
013	Fat, Acid Pretreat (%)	0202	7.287	0.4200	6.664	0.8099	0.1818	57	0.77	013.00	0
013	Fat, Acid Pretreat (%)	2150	7.290	0.3200	6.664	0.8099	0.1818	57	0.77	013.00	0
013	Fat, Acid Pretreat (%)	0553	7.325	0.0100	6.664	0.8099	0.1818	57	0.82	013.02	0
013	Fat, Acid Pretreat (%)	0682	7.340	0.0000	6.664	0.8099	0.1818	57	0.83	013.02	0
013	Fat, Acid Pretreat (%)	0164	7.425	0.0700	6.664	0.8099	0.1818	57	0.94	013.02	0
013	Fat, Acid Pretreat (%)	0035	7.430	0.1800	6.664	0.8099	0.1818	57	0.95	013.13	0
013	Fat, Acid Pretreat (%)	0955	7.510	0.0400	6.664	0.8099	0.1818	57	1.04	013.13	0
013	Fat, Acid Pretreat (%)	0675	7.520	0.0200	6.664	0.8099	0.1818	57	1.06	013.02	0
013	Fat, Acid Pretreat (%)	0425	7.535	0.0500	6.664	0.8099	0.1818	57	1.08	013.02	0
013	Fat, Acid Pretreat (%)	0676	7.830	0.4200	6.664	0.8099	0.1818	57	1.44	013.02	0
013	Fat, Acid Pretreat (%)	0407	7.845	0.2920	6.664	0.8099	0.1818	57	1.46	013.13	0
013	Fat, Acid Pretreat (%)	0034	8.260	0.3200	6.664	0.8099	0.1818	57	1.97	013.13	0
013	Fat, Acid Pretreat (%)	0265	8.425	0.1300	6.664	0.8099	0.1818	57	2.17	013.13	0
013	Fat, Acid Pretreat (%)	2128	9.228	0.0274	6.664	0.8099	0.1818	57	3.17	013.12	0
013	Fat, Acid Pretreat (%)	0643	8.540	2.700	6.664	0.8099	0.1818	57	2.32	013.02	1
015	Aluminum (mg / kg (ppm))	0186	100.5	4.774	152.7	15.94	4.284	14	-3.28	015.52	0
015	Aluminum (mg / kg (ppm))	0510	138.0	2.000	152.7	15.94	4.284	14	-0.92	015.43	0
015	Aluminum (mg / kg (ppm))	0918	139.7	3.391	152.7	15.94	4.284	14	-0.82	015.43	0
015	Aluminum (mg / kg (ppm))	0049	140.9	4.330	152.7	15.94	4.284	14	-0.74	015.41	0
015	Aluminum (mg / kg (ppm))	0425	143.9	2.217	152.7	15.94	4.284	14	-0.55	015.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
015	Aluminum (mg / kg (ppm))	0171	147.7	0.1500	152.7	15.94	4.284	14	-0.31	015.41	0
015	Aluminum (mg / kg (ppm))	0407	149.4	2.074	152.7	15.94	4.284	14	-0.21	015.41	0
015	Aluminum (mg / kg (ppm))	0169	153.0	4.000	152.7	15.94	4.284	14	0.02	015.43	0
015	Aluminum (mg / kg (ppm))	0560	157.6	12.30	152.7	15.94	4.284	14	0.30	015.42	0
015	Aluminum (mg / kg (ppm))	0164	158.0	4.000	152.7	15.94	4.284	14	0.33	015.41	0
015	Aluminum (mg / kg (ppm))	0353	165.5	1.000	152.7	15.94	4.284	14	0.80	015.43	0
015	Aluminum (mg / kg (ppm))	0520	167.0	14.00	152.7	15.94	4.284	14	0.90	015.41	0
015	Aluminum (mg / kg (ppm))	0553	172.5	1.000	152.7	15.94	4.284	14	1.24	015.53	0
015	Aluminum (mg / kg (ppm))	0964	176.1	4.735	152.7	15.94	4.284	14	1.47	015.43	0
015	Aluminum (mg / kg (ppm))	0407	147.7	24.03	152.7	15.94	4.284	14	-0.32	015.43	1
017	Boron (mg / kg (ppm))	0171	13.25	0.7000	17.69	2.630	0.4695	16	-1.69	017.41	0
017	Boron (mg / kg (ppm))	0358	13.53	0.0800	17.69	2.630	0.4695	16	-1.58	017.41	0
017	Boron (mg / kg (ppm))	0229	13.88	0.2400	17.69	2.630	0.4695	16	-1.45	017.41	0
017	Boron (mg / kg (ppm))	0049	15.88	0.1900	17.69	2.630	0.4695	16	-0.69	017.41	0
017	Boron (mg / kg (ppm))	0405	16.20	2.410	17.69	2.630	0.4695	16	-0.57	017.42	0
017	Boron (mg / kg (ppm))	0407	17.45	0.0322	17.69	2.630	0.4695	16	-0.09	017.41	0
017	Boron (mg / kg (ppm))	0918	18.50	0.0100	17.69	2.630	0.4695	16	0.31	017.43	0
017	Boron (mg / kg (ppm))	0353	18.50	1.000	17.69	2.630	0.4695	16	0.31	017.43	0
017	Boron (mg / kg (ppm))	0693	18.62	0.5970	17.69	2.630	0.4695	16	0.35	017.42	0
017	Boron (mg / kg (ppm))	0045	18.70	0.0000	17.69	2.630	0.4695	16	0.38	017.42	0
017	Boron (mg / kg (ppm))	0226	18.75	0.2500	17.69	2.630	0.4695	16	0.40	017.41	0
017	Boron (mg / kg (ppm))	0407	18.97	0.6020	17.69	2.630	0.4695	16	0.49	017.43	0
017	Boron (mg / kg (ppm))	0294	19.27	0.0600	17.69	2.630	0.4695	16	0.60	017.42	0
017	Boron (mg / kg (ppm))	0560	19.87	1.140	17.69	2.630	0.4695	16	0.83	017.42	0
017	Boron (mg / kg (ppm))	0510	20.00	0.0000	17.69	2.630	0.4695	16	0.88	017.43	0
017	Boron (mg / kg (ppm))	0083	21.00	0.2000	17.69	2.630	0.4695	16	1.26	017.43	0
017	Boron (mg / kg (ppm))	0405	17.92	3.200	17.69	2.630	0.4695	16	0.09	017.44	1
019	Calcium (%)	2022	0.3350	0.0100	0.4205	0.0334	0.0124	117	-2.56	019.00	0
019	Calcium (%)	0897	0.3450	0.0100	0.4205	0.0334	0.0124	117	-2.26	019.00	0
019	Calcium (%)	0294	0.3450	0.0100	0.4205	0.0334	0.0124	117	-2.26	019.42	0
019	Calcium (%)	0175	0.3550	0.0100	0.4205	0.0334	0.0124	117	-1.96	019.31	0
019	Calcium (%)	2022	0.3550	0.0100	0.4205	0.0334	0.0124	117	-1.96	019.31	0
019	Calcium (%)	0199	0.3550	0.0100	0.4205	0.0334	0.0124	117	-1.96	019.53	0
019	Calcium (%)	0144	0.3600	0.0000	0.4205	0.0334	0.0124	117	-1.81	019.41	0
019	Calcium (%)	0038	0.3655	0.0050	0.4205	0.0334	0.0124	117	-1.65	019.43	0
019	Calcium (%)	0885	0.3672	0.0062	0.4205	0.0334	0.0124	117	-1.59	019.08	0
019	Calcium (%)	2248	0.3710	0.0020	0.4205	0.0334	0.0124	117	-1.48	019.31	0
019	Calcium (%)	0171	0.3800	0.0000	0.4205	0.0334	0.0124	117	-1.21	019.41	0
019	Calcium (%)	0169	0.3850	0.0100	0.4205	0.0334	0.0124	117	-1.06	019.32	0
019	Calcium (%)	0229	0.3870	0.0060	0.4205	0.0334	0.0124	117	-1.00	019.41	0
019	Calcium (%)	2259	0.3880	0.0220	0.4205	0.0334	0.0124	117	-0.97	019.41	0
019	Calcium (%)	0881	0.3895	0.0050	0.4205	0.0334	0.0124	117	-0.93	019.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	0265	0.3900	0.0000	0.4205	0.0334	0.0124	117	-0.91	019.42	0
019	Calcium (%)	0083	0.3900	0.0000	0.4205	0.0334	0.0124	117	-0.91	019.43	0
019	Calcium (%)	0610	0.3900	0.0000	0.4205	0.0334	0.0124	117	-0.91	019.43	0
019	Calcium (%)	0186	0.3925	0.0138	0.4205	0.0334	0.0124	117	-0.84	019.52	0
019	Calcium (%)	0563	0.3926	0.0088	0.4205	0.0334	0.0124	117	-0.83	019.31	0
019	Calcium (%)	0017	0.3945	0.0330	0.4205	0.0334	0.0124	117	-0.78	019.43	0
019	Calcium (%)	0918	0.3945	0.0010	0.4205	0.0334	0.0124	117	-0.78	019.43	0
019	Calcium (%)	0536	0.3950	0.0100	0.4205	0.0334	0.0124	117	-0.76	019.31	0
019	Calcium (%)	0948	0.3980	0.0180	0.4205	0.0334	0.0124	117	-0.67	019.31	0
019	Calcium (%)	2259	0.3980	0.0120	0.4205	0.0334	0.0124	117	-0.67	019.42	0
019	Calcium (%)	0154	0.3982	0.0022	0.4205	0.0334	0.0124	117	-0.67	019.52	0
019	Calcium (%)	0357	0.3989	0.0015	0.4205	0.0334	0.0124	117	-0.65	019.42	0
019	Calcium (%)	0674	0.4000	0.0000	0.4205	0.0334	0.0124	117	-0.61	019.31	0
019	Calcium (%)	0049	0.4000	0.0000	0.4205	0.0334	0.0124	117	-0.61	019.41	0
019	Calcium (%)	0298	0.4000	0.0200	0.4205	0.0334	0.0124	117	-0.61	019.41	0
019	Calcium (%)	0358	0.4000	0.0200	0.4205	0.0334	0.0124	117	-0.61	019.41	0
019	Calcium (%)	0366	0.4000	0.0400	0.4205	0.0334	0.0124	117	-0.61	019.42	0
019	Calcium (%)	0425	0.4000	0.0020	0.4205	0.0334	0.0124	117	-0.61	019.43	0
019	Calcium (%)	0405	0.4010	0.0260	0.4205	0.0334	0.0124	117	-0.58	019.41	0
019	Calcium (%)	0042	0.4010	0.0060	0.4205	0.0334	0.0124	117	-0.58	019.42	0
019	Calcium (%)	2129	0.4045	0.0230	0.4205	0.0334	0.0124	117	-0.48	019.42	0
019	Calcium (%)	0904	0.4050	0.0100	0.4205	0.0334	0.0124	117	-0.46	019.31	0
019	Calcium (%)	0148	0.4050	0.0100	0.4205	0.0334	0.0124	117	-0.46	019.41	0
019	Calcium (%)	0511	0.4050	0.0300	0.4205	0.0334	0.0124	117	-0.46	019.41	0
019	Calcium (%)	0045	0.4055	0.0130	0.4205	0.0334	0.0124	117	-0.45	019.42	0
019	Calcium (%)	0560	0.4068	0.0129	0.4205	0.0334	0.0124	117	-0.41	019.42	0
019	Calcium (%)	0572	0.4075	0.0170	0.4205	0.0334	0.0124	117	-0.39	019.53	0
019	Calcium (%)	2196	0.4080	0.0000	0.4205	0.0334	0.0124	117	-0.37	019.08	0
019	Calcium (%)	0164	0.4080	0.0060	0.4205	0.0334	0.0124	117	-0.37	019.41	0
019	Calcium (%)	2062	0.4086	0.0141	0.4205	0.0334	0.0124	117	-0.36	019.00	0
019	Calcium (%)	0263	0.4099	0.0074	0.4205	0.0334	0.0124	117	-0.32	019.42	0
019	Calcium (%)	2196	0.4100	0.0000	0.4205	0.0334	0.0124	117	-0.31	019.31	0
019	Calcium (%)	2246	0.4100	0.0000	0.4205	0.0334	0.0124	117	-0.31	019.31	0
019	Calcium (%)	2012	0.4100	0.0000	0.4205	0.0334	0.0124	117	-0.31	019.41	0
019	Calcium (%)	0242	0.4100	0.0000	0.4205	0.0334	0.0124	117	-0.31	019.99	0
019	Calcium (%)	0590	0.4100	0.0200	0.4205	0.0334	0.0124	117	-0.31	019.08	0
019	Calcium (%)	0074	0.4100	0.0120	0.4205	0.0334	0.0124	117	-0.31	019.41	0
019	Calcium (%)	0520	0.4100	0.0200	0.4205	0.0334	0.0124	117	-0.31	019.41	0
019	Calcium (%)	0874	0.4105	0.0010	0.4205	0.0334	0.0124	117	-0.30	019.31	0
019	Calcium (%)	0693	0.4105	0.0090	0.4205	0.0334	0.0124	117	-0.30	019.42	0
019	Calcium (%)	0139	0.4150	0.0020	0.4205	0.0334	0.0124	117	-0.16	019.41	0
019	Calcium (%)	0035	0.4150	0.0100	0.4205	0.0334	0.0124	117	-0.16	019.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	2089	0.4150	0.0100	0.4205	0.0334	0.0124	117	-0.16	019.43	0
019	Calcium (%)	0108	0.4150	0.0300	0.4205	0.0334	0.0124	117	-0.16	019.31	0
019	Calcium (%)	0553	0.4155	0.0330	0.4205	0.0334	0.0124	117	-0.15	019.53	0
019	Calcium (%)	0038	0.4185	0.0202	0.4205	0.0334	0.0124	117	-0.06	019.31	0
019	Calcium (%)	0098	0.4200	0.0000	0.4205	0.0334	0.0124	117	-0.01	019.41	0
019	Calcium (%)	0884	0.4200	0.0400	0.4205	0.0334	0.0124	117	-0.01	019.31	0
019	Calcium (%)	0407	0.4222	0.0025	0.4205	0.0334	0.0124	117	0.05	019.41	0
019	Calcium (%)	0964	0.4238	0.0025	0.4205	0.0334	0.0124	117	0.10	019.43	0
019	Calcium (%)	2018	0.4250	0.0100	0.4205	0.0334	0.0124	117	0.13	019.31	0
019	Calcium (%)	0003	0.4250	0.0300	0.4205	0.0334	0.0124	117	0.13	019.41	0
019	Calcium (%)	0035	0.4250	0.0100	0.4205	0.0334	0.0124	117	0.13	019.43	0
019	Calcium (%)	0098	0.4250	0.0100	0.4205	0.0334	0.0124	117	0.13	019.43	0
019	Calcium (%)	0675	0.4250	0.0300	0.4205	0.0334	0.0124	117	0.13	019.43	0
019	Calcium (%)	0590	0.4250	0.0300	0.4205	0.0334	0.0124	117	0.13	019.99	0
019	Calcium (%)	0202	0.4265	0.0090	0.4205	0.0334	0.0124	117	0.18	019.43	0
019	Calcium (%)	2193	0.4280	0.0040	0.4205	0.0334	0.0124	117	0.22	019.08	0
019	Calcium (%)	0968	0.4290	0.0020	0.4205	0.0334	0.0124	117	0.25	019.43	0
019	Calcium (%)	0870	0.4299	0.0137	0.4205	0.0334	0.0124	117	0.28	019.42	0
019	Calcium (%)	0868	0.4300	0.0200	0.4205	0.0334	0.0124	117	0.28	019.31	0
019	Calcium (%)	0226	0.4300	0.0000	0.4205	0.0334	0.0124	117	0.28	019.41	0
019	Calcium (%)	0089	0.4300	0.0000	0.4205	0.0334	0.0124	117	0.28	019.43	0
019	Calcium (%)	0510	0.4300	0.0200	0.4205	0.0334	0.0124	117	0.28	019.43	0
019	Calcium (%)	2190	0.4308	0.0202	0.4205	0.0334	0.0124	117	0.31	019.08	0
019	Calcium (%)	0598	0.4336	0.0010	0.4205	0.0334	0.0124	117	0.39	019.41	0
019	Calcium (%)	0689	0.4350	0.0300	0.4205	0.0334	0.0124	117	0.43	019.08	0
019	Calcium (%)	0939	0.4350	0.0100	0.4205	0.0334	0.0124	117	0.43	019.31	0
019	Calcium (%)	2198	0.4350	0.0100	0.4205	0.0334	0.0124	117	0.43	019.31	0
019	Calcium (%)	2097	0.4377	0.0018	0.4205	0.0334	0.0124	117	0.51	019.31	0
019	Calcium (%)	0619	0.4400	0.0160	0.4205	0.0334	0.0124	117	0.58	019.41	0
019	Calcium (%)	0910	0.4400	0.0200	0.4205	0.0334	0.0124	117	0.58	019.41	0
019	Calcium (%)	0682	0.4400	0.0000	0.4205	0.0334	0.0124	117	0.58	019.43	0
019	Calcium (%)	0939	0.4400	0.0200	0.4205	0.0334	0.0124	117	0.58	019.53	0
019	Calcium (%)	0100	0.4400	0.0000	0.4205	0.0334	0.0124	117	0.58	019.99	0
019	Calcium (%)	0186	0.4442	0.0193	0.4205	0.0334	0.0124	117	0.71	019.42	0
019	Calcium (%)	0622	0.4444	0.0634	0.4205	0.0334	0.0124	117	0.72	019.31	0
019	Calcium (%)	0651	0.4455	0.0010	0.4205	0.0334	0.0124	117	0.75	019.00	0
019	Calcium (%)	2109	0.4465	0.0230	0.4205	0.0334	0.0124	117	0.78	019.00	0
019	Calcium (%)	0720	0.4500	0.0000	0.4205	0.0334	0.0124	117	0.88	019.41	0
019	Calcium (%)	0676	0.4500	0.0200	0.4205	0.0334	0.0124	117	0.88	019.99	0
019	Calcium (%)	0037	0.4520	0.0080	0.4205	0.0334	0.0124	117	0.94	019.42	0
019	Calcium (%)	0278	0.4550	0.0500	0.4205	0.0334	0.0124	117	1.03	019.42	0
019	Calcium (%)	0505	0.4600	0.0200	0.4205	0.0334	0.0124	117	1.18	019.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	0683	0.4600	0.0000	0.4205	0.0334	0.0124	117	1.18	019.00	0
019	Calcium (%)	0529	0.4650	0.0100	0.4205	0.0334	0.0124	117	1.33	019.31	0
019	Calcium (%)	0142	0.4663	0.0196	0.4205	0.0334	0.0124	117	1.37	019.31	0
019	Calcium (%)	0407	0.4687	0.0091	0.4205	0.0334	0.0124	117	1.44	019.43	0
019	Calcium (%)	0190	0.4700	0.0200	0.4205	0.0334	0.0124	117	1.48	019.42	0
019	Calcium (%)	0353	0.4700	0.0200	0.4205	0.0334	0.0124	117	1.48	019.43	0
019	Calcium (%)	0726	0.4760	0.0018	0.4205	0.0334	0.0124	117	1.66	019.42	0
019	Calcium (%)	0032	0.4775	0.0409	0.4205	0.0334	0.0124	117	1.70	019.43	0
019	Calcium (%)	0849	0.4775	0.0150	0.4205	0.0334	0.0124	117	1.71	019.08	0
019	Calcium (%)	2188	0.4800	0.0200	0.4205	0.0334	0.0124	117	1.78	019.08	0
019	Calcium (%)	2251	0.4800	0.0000	0.4205	0.0334	0.0124	117	1.78	019.99	0
019	Calcium (%)	0354	0.5050	0.0100	0.4205	0.0334	0.0124	117	2.53	019.41	0
019	Calcium (%)	0036	0.5053	0.0097	0.4205	0.0334	0.0124	117	2.54	019.03	0
019	Calcium (%)	0685	0.5300	0.0200	0.4205	0.0334	0.0124	117	3.28	019.41	0
019	Calcium (%)	2128	0.5470	0.0020	0.4205	0.0334	0.0124	117	3.79	019.02	0
019	Calcium (%)	0405	0.6230	0.0120	0.4205	0.0334	0.0124	117	6.06	019.42	0
019	Calcium (%)	2006	0.6790	0.0280	0.4205	0.0334	0.0124	117	7.73	019.00	0
019	Calcium (%)	0852	1.000	0.0000	0.4205	0.0334	0.0124	117	17.34	019.08	0
019	Calcium (%)	0687	0.4000	0.1000	0.4205	0.0334	0.0124	117	-0.61	019.31	1
019	Calcium (%)	2192	0.5450	0.1300	0.4205	0.0334	0.0124	117	3.73	019.43	1
019	Calcium (%)	0536	2.100	0.1600	0.4205	0.0334	0.0124	117	50.25	019.02	2
019	Calcium (%)	2031	2.588	0.0489	0.4205	0.0334	0.0124	117	64.86	019.08	2
021	Cobalt (mg / kg (ppm))	0619	0.0560	0.0000	0.9556	0.4906	0.0759	22	-1.83	021.41	0
021	Cobalt (mg / kg (ppm))	0910	0.2850	0.0500	0.9556	0.4906	0.0759	22	-1.37	021.52	0
021	Cobalt (mg / kg (ppm))	0918	0.3034	0.0094	0.9556	0.4906	0.0759	22	-1.33	021.53	0
021	Cobalt (mg / kg (ppm))	0186	0.3310	0.0080	0.9556	0.4906	0.0759	22	-1.27	021.52	0
021	Cobalt (mg / kg (ppm))	0047	0.3600	0.0000	0.9556	0.4906	0.0759	22	-1.21	021.52	0
021	Cobalt (mg / kg (ppm))	0199	0.5550	0.0100	0.9556	0.4906	0.0759	22	-0.82	021.53	0
021	Cobalt (mg / kg (ppm))	0693	0.8210	0.1460	0.9556	0.4906	0.0759	22	-0.27	021.42	0
021	Cobalt (mg / kg (ppm))	0560	0.9100	0.1800	0.9556	0.4906	0.0759	22	-0.09	021.42	0
021	Cobalt (mg / kg (ppm))	0553	0.9300	0.0740	0.9556	0.4906	0.0759	22	-0.05	021.53	0
021	Cobalt (mg / kg (ppm))	0964	0.9838	0.0856	0.9556	0.4906	0.0759	22	0.06	021.43	0
021	Cobalt (mg / kg (ppm))	0510	1.060	0.0200	0.9556	0.4906	0.0759	22	0.21	021.43	0
021	Cobalt (mg / kg (ppm))	2248	1.080	0.1400	0.9556	0.4906	0.0759	22	0.25	021.31	0
021	Cobalt (mg / kg (ppm))	0171	1.130	0.1200	0.9556	0.4906	0.0759	22	0.36	021.41	0
021	Cobalt (mg / kg (ppm))	0610	1.130	0.1200	0.9556	0.4906	0.0759	22	0.36	021.53	0
021	Cobalt (mg / kg (ppm))	0939	1.195	0.0700	0.9556	0.4906	0.0759	22	0.49	021.31	0
021	Cobalt (mg / kg (ppm))	0939	1.235	0.0500	0.9556	0.4906	0.0759	22	0.57	021.53	0
021	Cobalt (mg / kg (ppm))	0407	1.273	0.0575	0.9556	0.4906	0.0759	22	0.65	021.41	0
021	Cobalt (mg / kg (ppm))	0164	1.300	0.0000	0.9556	0.4906	0.0759	22	0.70	021.31	0
021	Cobalt (mg / kg (ppm))	0169	1.315	0.1300	0.9556	0.4906	0.0759	22	0.73	021.43	0
021	Cobalt (mg / kg (ppm))	0148	1.500	0.0400	0.9556	0.4906	0.0759	22	1.11	021.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
021	Cobalt (mg / kg (ppm))	0038	1.505	0.0160	0.9556	0.4906	0.0759	22	1.12	021.43	0
021	Cobalt (mg / kg (ppm))	0407	1.602	0.3423	0.9556	0.4906	0.0759	22	1.32	021.43	0
021	Cobalt (mg / kg (ppm))	0045	1.119	0.5230	0.9556	0.4906	0.0759	22	0.33	021.42	1
022	Copper (mg / kg (ppm))	2188	7.875	0.1300	13.14	1.263	0.6443	89	-4.17	022.31	0
022	Copper (mg / kg (ppm))	0536	8.525	0.1500	13.14	1.263	0.6443	89	-3.65	022.31	0
022	Copper (mg / kg (ppm))	0675	8.820	1.400	13.14	1.263	0.6443	89	-3.42	022.43	0
022	Copper (mg / kg (ppm))	0017	10.76	0.9940	13.14	1.263	0.6443	89	-1.88	022.43	0
022	Copper (mg / kg (ppm))	0529	10.90	2.000	13.14	1.263	0.6443	89	-1.77	022.31	0
022	Copper (mg / kg (ppm))	0511	11.00	0.0000	13.14	1.263	0.6443	89	-1.69	022.41	0
022	Copper (mg / kg (ppm))	0242	11.00	0.0000	13.14	1.263	0.6443	89	-1.69	022.99	0
022	Copper (mg / kg (ppm))	0045	11.30	1.400	13.14	1.263	0.6443	89	-1.46	022.42	0
022	Copper (mg / kg (ppm))	0939	11.44	0.1500	13.14	1.263	0.6443	89	-1.35	022.31	0
022	Copper (mg / kg (ppm))	0100	11.50	1.000	13.14	1.263	0.6443	89	-1.30	022.99	0
022	Copper (mg / kg (ppm))	2129	11.55	0.5000	13.14	1.263	0.6443	89	-1.26	022.99	0
022	Copper (mg / kg (ppm))	0169	11.60	0.4000	13.14	1.263	0.6443	89	-1.22	022.43	0
022	Copper (mg / kg (ppm))	0939	11.67	0.5600	13.14	1.263	0.6443	89	-1.16	022.53	0
022	Copper (mg / kg (ppm))	0848	11.78	0.0300	13.14	1.263	0.6443	89	-1.08	022.41	0
022	Copper (mg / kg (ppm))	0263	11.98	0.2566	13.14	1.263	0.6443	89	-0.91	022.42	0
022	Copper (mg / kg (ppm))	0074	12.00	4.000	13.14	1.263	0.6443	89	-0.90	022.41	0
022	Copper (mg / kg (ppm))	0186	12.00	0.0000	13.14	1.263	0.6443	89	-0.90	022.42	0
022	Copper (mg / kg (ppm))	0366	12.00	0.0000	13.14	1.263	0.6443	89	-0.90	022.42	0
022	Copper (mg / kg (ppm))	0353	12.00	2.000	13.14	1.263	0.6443	89	-0.90	022.43	0
022	Copper (mg / kg (ppm))	0186	12.15	0.5000	13.14	1.263	0.6443	89	-0.78	022.52	0
022	Copper (mg / kg (ppm))	0278	12.25	0.1000	13.14	1.263	0.6443	89	-0.70	022.42	0
022	Copper (mg / kg (ppm))	0918	12.33	0.0400	13.14	1.263	0.6443	89	-0.64	022.43	0
022	Copper (mg / kg (ppm))	0553	12.35	0.1000	13.14	1.263	0.6443	89	-0.62	022.53	0
022	Copper (mg / kg (ppm))	0520	12.50	1.000	13.14	1.263	0.6443	89	-0.51	022.41	0
022	Copper (mg / kg (ppm))	0910	12.50	1.000	13.14	1.263	0.6443	89	-0.51	022.41	0
022	Copper (mg / kg (ppm))	0407	12.58	0.1142	13.14	1.263	0.6443	89	-0.44	022.41	0
022	Copper (mg / kg (ppm))	0682	12.60	0.0000	13.14	1.263	0.6443	89	-0.43	022.43	0
022	Copper (mg / kg (ppm))	0884	12.65	0.5000	13.14	1.263	0.6443	89	-0.39	022.31	0
022	Copper (mg / kg (ppm))	0049	12.69	0.9600	13.14	1.263	0.6443	89	-0.36	022.41	0
022	Copper (mg / kg (ppm))	0026	12.79	0.3700	13.14	1.263	0.6443	89	-0.27	022.42	0
022	Copper (mg / kg (ppm))	0038	12.80	1.400	13.14	1.263	0.6443	89	-0.27	022.32	0
022	Copper (mg / kg (ppm))	0870	12.80	0.3600	13.14	1.263	0.6443	89	-0.27	022.42	0
022	Copper (mg / kg (ppm))	0199	12.80	0.1800	13.14	1.263	0.6443	89	-0.27	022.53	0
022	Copper (mg / kg (ppm))	0358	12.89	0.5400	13.14	1.263	0.6443	89	-0.20	022.41	0
022	Copper (mg / kg (ppm))	0726	12.93	0.0200	13.14	1.263	0.6443	89	-0.17	022.42	0
022	Copper (mg / kg (ppm))	2248	12.95	1.100	13.14	1.263	0.6443	89	-0.15	022.31	0
022	Copper (mg / kg (ppm))	0098	12.99	0.0200	13.14	1.263	0.6443	89	-0.12	022.43	0
022	Copper (mg / kg (ppm))	0164	13.00	0.0000	13.14	1.263	0.6443	89	-0.11	022.41	0
022	Copper (mg / kg (ppm))	0294	13.00	0.2000	13.14	1.263	0.6443	89	-0.11	022.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
022	Copper (mg / kg (ppm))	0035	13.00	0.0000	13.14	1.263	0.6443	89	-0.11	022.43	0
022	Copper (mg / kg (ppm))	0083	13.00	0.0000	13.14	1.263	0.6443	89	-0.11	022.43	0
022	Copper (mg / kg (ppm))	0510	13.00	0.0000	13.14	1.263	0.6443	89	-0.11	022.43	0
022	Copper (mg / kg (ppm))	0968	13.00	0.0000	13.14	1.263	0.6443	89	-0.11	022.43	0
022	Copper (mg / kg (ppm))	0590	13.00	0.8000	13.14	1.263	0.6443	89	-0.11	022.99	0
022	Copper (mg / kg (ppm))	0265	13.05	0.3000	13.14	1.263	0.6443	89	-0.07	022.42	0
022	Copper (mg / kg (ppm))	0572	13.10	0.8000	13.14	1.263	0.6443	89	-0.03	022.53	0
022	Copper (mg / kg (ppm))	0032	13.13	0.6900	13.14	1.263	0.6443	89	-0.01	022.43	0
022	Copper (mg / kg (ppm))	0693	13.15	0.2570	13.14	1.263	0.6443	89	0.00	022.42	0
022	Copper (mg / kg (ppm))	0904	13.17	0.0200	13.14	1.263	0.6443	89	0.02	022.31	0
022	Copper (mg / kg (ppm))	0407	13.25	0.4323	13.14	1.263	0.6443	89	0.09	022.43	0
022	Copper (mg / kg (ppm))	0964	13.26	0.0174	13.14	1.263	0.6443	89	0.09	022.43	0
022	Copper (mg / kg (ppm))	0004	13.31	1.280	13.14	1.263	0.6443	89	0.14	022.41	0
022	Copper (mg / kg (ppm))	0010	13.40	0.2000	13.14	1.263	0.6443	89	0.21	022.33	0
022	Copper (mg / kg (ppm))	0229	13.49	1.070	13.14	1.263	0.6443	89	0.27	022.41	0
022	Copper (mg / kg (ppm))	0685	13.52	0.0300	13.14	1.263	0.6443	89	0.30	022.42	0
022	Copper (mg / kg (ppm))	0560	13.55	0.5000	13.14	1.263	0.6443	89	0.33	022.42	0
022	Copper (mg / kg (ppm))	0202	13.58	0.5690	13.14	1.263	0.6443	89	0.35	022.43	0
022	Copper (mg / kg (ppm))	2097	13.60	0.0000	13.14	1.263	0.6443	89	0.36	022.31	0
022	Copper (mg / kg (ppm))	2146	13.62	0.3600	13.14	1.263	0.6443	89	0.38	022.31	0
022	Copper (mg / kg (ppm))	0038	13.65	0.1000	13.14	1.263	0.6443	89	0.40	022.43	0
022	Copper (mg / kg (ppm))	0598	13.71	0.2600	13.14	1.263	0.6443	89	0.45	022.41	0
022	Copper (mg / kg (ppm))	0425	13.75	0.7380	13.14	1.263	0.6443	89	0.48	022.43	0
022	Copper (mg / kg (ppm))	0948	13.76	0.7850	13.14	1.263	0.6443	89	0.49	022.33	0
022	Copper (mg / kg (ppm))	2089	13.78	2.230	13.14	1.263	0.6443	89	0.50	022.43	0
022	Copper (mg / kg (ppm))	2062	13.90	0.5980	13.14	1.263	0.6443	89	0.60	022.31	0
022	Copper (mg / kg (ppm))	0190	13.94	0.0600	13.14	1.263	0.6443	89	0.63	022.42	0
022	Copper (mg / kg (ppm))	0357	13.95	0.0210	13.14	1.263	0.6443	89	0.64	022.42	0
022	Copper (mg / kg (ppm))	0098	13.99	0.5000	13.14	1.263	0.6443	89	0.67	022.41	0
022	Copper (mg / kg (ppm))	0035	14.00	0.0000	13.14	1.263	0.6443	89	0.68	022.42	0
022	Copper (mg / kg (ppm))	2146	14.03	0.2000	13.14	1.263	0.6443	89	0.71	022.51	0
022	Copper (mg / kg (ppm))	0148	14.08	0.1300	13.14	1.263	0.6443	89	0.74	022.41	0
022	Copper (mg / kg (ppm))	0037	14.15	0.7000	13.14	1.263	0.6443	89	0.80	022.42	0
022	Copper (mg / kg (ppm))	0009	14.20	5.400	13.14	1.263	0.6443	89	0.84	022.42	0
022	Copper (mg / kg (ppm))	0505	14.25	0.0300	13.14	1.263	0.6443	89	0.88	022.43	0
022	Copper (mg / kg (ppm))	0027	14.27	0.6780	13.14	1.263	0.6443	89	0.90	022.43	0
022	Copper (mg / kg (ppm))	0619	14.45	1.700	13.14	1.263	0.6443	89	1.04	022.41	0
022	Copper (mg / kg (ppm))	0047	14.47	0.3300	13.14	1.263	0.6443	89	1.05	022.52	0
022	Copper (mg / kg (ppm))	2196	14.50	0.0000	13.14	1.263	0.6443	89	1.08	022.31	0
022	Copper (mg / kg (ppm))	0171	14.50	0.6000	13.14	1.263	0.6443	89	1.08	022.41	0
022	Copper (mg / kg (ppm))	0226	14.61	0.9300	13.14	1.263	0.6443	89	1.16	022.41	0
022	Copper (mg / kg (ppm))	0042	15.00	2.200	13.14	1.263	0.6443	89	1.47	022.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
022	Copper (mg / kg (ppm))	0610	15.00	0.0000	13.14	1.263	0.6443	89	1.47	022.43	0
022	Copper (mg / kg (ppm))	0175	15.50	1.000	13.14	1.263	0.6443	89	1.87	022.31	0
022	Copper (mg / kg (ppm))	0354	15.67	0.0700	13.14	1.263	0.6443	89	2.00	022.41	0
022	Copper (mg / kg (ppm))	0868	16.00	0.0000	13.14	1.263	0.6443	89	2.26	022.31	0
022	Copper (mg / kg (ppm))	2192	16.08	5.330	13.14	1.263	0.6443	89	2.32	022.43	0
022	Copper (mg / kg (ppm))	0066	17.98	0.4100	13.14	1.263	0.6443	89	3.83	022.44	0
022	Copper (mg / kg (ppm))	2012	18.69	0.1900	13.14	1.263	0.6443	89	4.39	022.41	0
022	Copper (mg / kg (ppm))	0405	27.14	1.350	13.14	1.263	0.6443	89	11.08	022.41	0
022	Copper (mg / kg (ppm))	2128	9.597	8.575	13.14	1.263	0.6443	89	-2.80	022.32	1
022	Copper (mg / kg (ppm))	0003	23.00	14.00	13.14	1.263	0.6443	89	7.81	022.41	1
022	Copper (mg / kg (ppm))	0405	69.12	2.930	13.14	1.263	0.6443	89	44.31	022.42	2
022	Copper (mg / kg (ppm))	0720	0.0000	0.0000	13.14	1.263	0.6443	89		022.41	4
024	Iodine (mg / kg (ppm))	0160	0.3500	0.0600				1		024.99	0
025	Iron (mg / kg (ppm))	0047	133.4	4.900	257.0	28.41	8.912	84	-4.35	025.52	0
025	Iron (mg / kg (ppm))	2248	169.7	5.060	257.0	28.41	8.912	84	-3.07	025.31	0
025	Iron (mg / kg (ppm))	0939	194.9	2.860	257.0	28.41	8.912	84	-2.19	025.31	0
025	Iron (mg / kg (ppm))	0294	197.2	4.260	257.0	28.41	8.912	84	-2.11	025.42	0
025	Iron (mg / kg (ppm))	0610	206.5	5.000	257.0	28.41	8.912	84	-1.78	025.43	0
025	Iron (mg / kg (ppm))	0265	208.0	2.000	257.0	28.41	8.912	84	-1.73	025.42	0
025	Iron (mg / kg (ppm))	0598	211.2	1.000	257.0	28.41	8.912	84	-1.61	025.41	0
025	Iron (mg / kg (ppm))	0199	212.7	12.25	257.0	28.41	8.912	84	-1.56	025.53	0
025	Iron (mg / kg (ppm))	0425	213.7	3.800	257.0	28.41	8.912	84	-1.52	025.43	0
025	Iron (mg / kg (ppm))	0726	221.6	0.2000	257.0	28.41	8.912	84	-1.25	025.42	0
025	Iron (mg / kg (ppm))	0278	222.0	14.00	257.0	28.41	8.912	84	-1.23	025.42	0
025	Iron (mg / kg (ppm))	0720	222.7	1.700	257.0	28.41	8.912	84	-1.21	025.41	0
025	Iron (mg / kg (ppm))	2146	223.7	3.750	257.0	28.41	8.912	84	-1.17	025.31	0
025	Iron (mg / kg (ppm))	0354	225.4	2.850	257.0	28.41	8.912	84	-1.11	025.41	0
025	Iron (mg / kg (ppm))	0939	233.5	3.000	257.0	28.41	8.912	84	-0.83	025.53	0
025	Iron (mg / kg (ppm))	0874	235.0	12.00	257.0	28.41	8.912	84	-0.78	025.31	0
025	Iron (mg / kg (ppm))	0553	235.5	31.00	257.0	28.41	8.912	84	-0.76	025.53	0
025	Iron (mg / kg (ppm))	0529	237.7	2.400	257.0	28.41	8.912	84	-0.68	025.31	0
025	Iron (mg / kg (ppm))	0190	238.4	1.010	257.0	28.41	8.912	84	-0.66	025.42	0
025	Iron (mg / kg (ppm))	0619	239.0	4.000	257.0	28.41	8.912	84	-0.64	025.41	0
025	Iron (mg / kg (ppm))	0560	240.8	28.10	257.0	28.41	8.912	84	-0.57	025.42	0
025	Iron (mg / kg (ppm))	0003	241.0	28.00	257.0	28.41	8.912	84	-0.56	025.41	0
025	Iron (mg / kg (ppm))	0154	242.1	8.320	257.0	28.41	8.912	84	-0.53	025.52	0
025	Iron (mg / kg (ppm))	0226	242.7	12.83	257.0	28.41	8.912	84	-0.51	025.41	0
025	Iron (mg / kg (ppm))	0675	243.2	3.890	257.0	28.41	8.912	84	-0.49	025.43	0
025	Iron (mg / kg (ppm))	0918	245.5	5.710	257.0	28.41	8.912	84	-0.41	025.43	0
025	Iron (mg / kg (ppm))	2196	245.6	0.0000	257.0	28.41	8.912	84	-0.40	025.31	0
025	Iron (mg / kg (ppm))	0511	247.5	13.00	257.0	28.41	8.912	84	-0.34	025.41	0
025	Iron (mg / kg (ppm))	2192	247.8	3.700	257.0	28.41	8.912	84	-0.32	025.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
025	Iron (mg / kg (ppm))	0175	248.0	0.0000	257.0	28.41	8.912	84	-0.32	025.31	0
025	Iron (mg / kg (ppm))	0026	249.8	0.5300	257.0	28.41	8.912	84	-0.26	025.42	0
025	Iron (mg / kg (ppm))	0083	250.5	1.000	257.0	28.41	8.912	84	-0.23	025.43	0
025	Iron (mg / kg (ppm))	2022	251.9	6.514	257.0	28.41	8.912	84	-0.18	025.31	0
025	Iron (mg / kg (ppm))	2129	253.0	12.00	257.0	28.41	8.912	84	-0.14	025.99	0
025	Iron (mg / kg (ppm))	0520	253.5	9.000	257.0	28.41	8.912	84	-0.12	025.41	0
025	Iron (mg / kg (ppm))	0968	253.5	1.000	257.0	28.41	8.912	84	-0.12	025.43	0
025	Iron (mg / kg (ppm))	0263	254.2	1.310	257.0	28.41	8.912	84	-0.10	025.42	0
025	Iron (mg / kg (ppm))	2012	254.6	8.710	257.0	28.41	8.912	84	-0.09	025.41	0
025	Iron (mg / kg (ppm))	0004	255.0	7.370	257.0	28.41	8.912	84	-0.07	025.41	0
025	Iron (mg / kg (ppm))	0169	255.0	12.00	257.0	28.41	8.912	84	-0.07	025.43	0
025	Iron (mg / kg (ppm))	0074	255.5	5.000	257.0	28.41	8.912	84	-0.05	025.41	0
025	Iron (mg / kg (ppm))	0164	257.0	8.000	257.0	28.41	8.912	84	0.00	025.41	0
025	Iron (mg / kg (ppm))	0242	257.0	2.000	257.0	28.41	8.912	84	0.00	025.99	0
025	Iron (mg / kg (ppm))	0049	258.9	0.3600	257.0	28.41	8.912	84	0.07	025.41	0
025	Iron (mg / kg (ppm))	0171	258.9	0.6000	257.0	28.41	8.912	84	0.07	025.41	0
025	Iron (mg / kg (ppm))	0001	259.4	0.4000	257.0	28.41	8.912	84	0.08	025.31	0
025	Iron (mg / kg (ppm))	0405	260.0	21.90	257.0	28.41	8.912	84	0.10	025.41	0
025	Iron (mg / kg (ppm))	2146	260.0	3.450	257.0	28.41	8.912	84	0.11	025.51	0
025	Iron (mg / kg (ppm))	0407	260.4	2.640	257.0	28.41	8.912	84	0.12	025.41	0
025	Iron (mg / kg (ppm))	0405	260.4	3.880	257.0	28.41	8.912	84	0.12	025.42	0
025	Iron (mg / kg (ppm))	0017	260.5	2.118	257.0	28.41	8.912	84	0.12	025.43	0
025	Iron (mg / kg (ppm))	0148	263.4	7.670	257.0	28.41	8.912	84	0.22	025.41	0
025	Iron (mg / kg (ppm))	0868	265.0	2.000	257.0	28.41	8.912	84	0.28	025.31	0
025	Iron (mg / kg (ppm))	0510	266.0	2.000	257.0	28.41	8.912	84	0.32	025.43	0
025	Iron (mg / kg (ppm))	0366	267.0	0.0000	257.0	28.41	8.912	84	0.35	025.42	0
025	Iron (mg / kg (ppm))	0038	268.0	4.000	257.0	28.41	8.912	84	0.39	025.43	0
025	Iron (mg / kg (ppm))	0045	268.5	3.000	257.0	28.41	8.912	84	0.40	025.42	0
025	Iron (mg / kg (ppm))	0358	268.5	7.600	257.0	28.41	8.912	84	0.40	025.41	0
025	Iron (mg / kg (ppm))	0693	269.3	30.68	257.0	28.41	8.912	84	0.43	025.42	0
025	Iron (mg / kg (ppm))	0035	270.0	10.00	257.0	28.41	8.912	84	0.46	025.42	0
025	Iron (mg / kg (ppm))	0038	273.5	5.000	257.0	28.41	8.912	84	0.58	025.31	0
025	Iron (mg / kg (ppm))	0353	274.0	8.000	257.0	28.41	8.912	84	0.60	025.43	0
025	Iron (mg / kg (ppm))	0035	276.5	7.000	257.0	28.41	8.912	84	0.68	025.43	0
025	Iron (mg / kg (ppm))	0229	277.5	4.300	257.0	28.41	8.912	84	0.72	025.41	0
025	Iron (mg / kg (ppm))	0100	277.5	25.00	257.0	28.41	8.912	84	0.72	025.99	0
025	Iron (mg / kg (ppm))	0042	278.0	2.000	257.0	28.41	8.912	84	0.74	025.42	0
025	Iron (mg / kg (ppm))	0098	278.3	8.700	257.0	28.41	8.912	84	0.75	025.43	0
025	Iron (mg / kg (ppm))	0964	278.4	16.20	257.0	28.41	8.912	84	0.75	025.43	0
025	Iron (mg / kg (ppm))	0098	279.9	29.90	257.0	28.41	8.912	84	0.80	025.41	0
025	Iron (mg / kg (ppm))	0066	281.9	6.380	257.0	28.41	8.912	84	0.87	025.41	0
025	Iron (mg / kg (ppm))	0407	282.1	2.294	257.0	28.41	8.912	84	0.88	025.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
025	Iron (mg / kg (ppm))	0027	284.1	11.37	257.0	28.41	8.912	84	0.95	025.43	0
025	Iron (mg / kg (ppm))	0682	285.0	0.0000	257.0	28.41	8.912	84	0.98	025.43	0
025	Iron (mg / kg (ppm))	0910	289.5	19.00	257.0	28.41	8.912	84	1.14	025.41	0
025	Iron (mg / kg (ppm))	0037	293.1	22.10	257.0	28.41	8.912	84	1.27	025.42	0
025	Iron (mg / kg (ppm))	2097	294.3	3.670	257.0	28.41	8.912	84	1.31	025.31	0
025	Iron (mg / kg (ppm))	0884	298.0	14.40	257.0	28.41	8.912	84	1.44	025.31	0
025	Iron (mg / kg (ppm))	0032	309.0	53.80	257.0	28.41	8.912	84	1.83	025.43	0
025	Iron (mg / kg (ppm))	0904	309.4	17.91	257.0	28.41	8.912	84	1.84	025.31	0
025	Iron (mg / kg (ppm))	0870	315.4	10.70	257.0	28.41	8.912	84	2.05	025.42	0
025	Iron (mg / kg (ppm))	2259	323.5	20.99	257.0	28.41	8.912	84	2.34	025.42	0
025	Iron (mg / kg (ppm))	2188	445.7	20.48	257.0	28.41	8.912	84	6.64	025.31	0
025	Iron (mg / kg (ppm))	2246	499.0	4.000	257.0	28.41	8.912	84	8.52	025.31	0
025	Iron (mg / kg (ppm))	0563	714.5	38.09	257.0	28.41	8.912	84	16.10	025.31	0
025	Iron (mg / kg (ppm))	2128	156.1	153.8	257.0	28.41	8.912	84	-3.55	025.32	1
025	Iron (mg / kg (ppm))	0536	1,436	19.69	257.0	28.41	8.912	84	41.50	025.31	2
027	Magnesium (%)	0169	0.3200	0.0200	0.3805	0.0207	0.0059	73	-2.93	027.32	0
027	Magnesium (%)	0563	0.3311	0.0027	0.3805	0.0207	0.0059	73	-2.39	027.31	0
027	Magnesium (%)	0038	0.3367	0.0050	0.3805	0.0207	0.0059	73	-2.12	027.31	0
027	Magnesium (%)	0089	0.3400	0.0000	0.3805	0.0207	0.0059	73	-1.96	027.43	0
027	Magnesium (%)	2129	0.3440	0.0040	0.3805	0.0207	0.0059	73	-1.77	027.42	0
027	Magnesium (%)	0035	0.3500	0.0000	0.3805	0.0207	0.0059	73	-1.48	027.43	0
027	Magnesium (%)	0511	0.3550	0.0100	0.3805	0.0207	0.0059	73	-1.23	027.41	0
027	Magnesium (%)	0407	0.3569	0.0126	0.3805	0.0207	0.0059	73	-1.14	027.41	0
027	Magnesium (%)	0354	0.3600	0.0000	0.3805	0.0207	0.0059	73	-0.99	027.41	0
027	Magnesium (%)	0358	0.3600	0.0000	0.3805	0.0207	0.0059	73	-0.99	027.41	0
027	Magnesium (%)	0294	0.3600	0.0000	0.3805	0.0207	0.0059	73	-0.99	027.42	0
027	Magnesium (%)	0083	0.3600	0.0000	0.3805	0.0207	0.0059	73	-0.99	027.43	0
027	Magnesium (%)	0353	0.3600	0.0200	0.3805	0.0207	0.0059	73	-0.99	027.43	0
027	Magnesium (%)	0610	0.3600	0.0000	0.3805	0.0207	0.0059	73	-0.99	027.43	0
027	Magnesium (%)	0186	0.3601	0.0077	0.3805	0.0207	0.0059	73	-0.99	027.52	0
027	Magnesium (%)	0918	0.3630	0.0020	0.3805	0.0207	0.0059	73	-0.85	027.43	0
027	Magnesium (%)	0263	0.3642	0.0020	0.3805	0.0207	0.0059	73	-0.79	027.42	0
027	Magnesium (%)	0536	0.3650	0.0100	0.3805	0.0207	0.0059	73	-0.75	027.31	0
027	Magnesium (%)	0035	0.3650	0.0100	0.3805	0.0207	0.0059	73	-0.75	027.42	0
027	Magnesium (%)	0202	0.3660	0.0040	0.3805	0.0207	0.0059	73	-0.70	027.43	0
027	Magnesium (%)	0038	0.3690	0.0000	0.3805	0.0207	0.0059	73	-0.56	027.43	0
027	Magnesium (%)	2012	0.3700	0.0000	0.3805	0.0207	0.0059	73	-0.51	027.41	0
027	Magnesium (%)	0265	0.3700	0.0000	0.3805	0.0207	0.0059	73	-0.51	027.42	0
027	Magnesium (%)	0425	0.3700	0.0000	0.3805	0.0207	0.0059	73	-0.51	027.43	0
027	Magnesium (%)	0229	0.3740	0.0100	0.3805	0.0207	0.0059	73	-0.32	027.41	0
027	Magnesium (%)	0154	0.3743	0.0123	0.3805	0.0207	0.0059	73	-0.30	027.52	0
027	Magnesium (%)	0874	0.3750	0.0000	0.3805	0.0207	0.0059	73	-0.27	027.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
027	Magnesium (%)	0884	0.3750	0.0100	0.3805	0.0207	0.0059	73	-0.27	027.31	0
027	Magnesium (%)	0049	0.3750	0.0100	0.3805	0.0207	0.0059	73	-0.27	027.41	0
027	Magnesium (%)	0139	0.3785	0.0004	0.3805	0.0207	0.0059	73	-0.10	027.41	0
027	Magnesium (%)	0074	0.3790	0.0140	0.3805	0.0207	0.0059	73	-0.07	027.41	0
027	Magnesium (%)	0868	0.3800	0.0000	0.3805	0.0207	0.0059	73	-0.03	027.31	0
027	Magnesium (%)	0003	0.3800	0.0000	0.3805	0.0207	0.0059	73	-0.03	027.41	0
027	Magnesium (%)	0148	0.3800	0.0000	0.3805	0.0207	0.0059	73	-0.03	027.41	0
027	Magnesium (%)	0720	0.3800	0.0000	0.3805	0.0207	0.0059	73	-0.03	027.41	0
027	Magnesium (%)	0045	0.3800	0.0140	0.3805	0.0207	0.0059	73	-0.03	027.42	0
027	Magnesium (%)	0278	0.3800	0.0000	0.3805	0.0207	0.0059	73	-0.03	027.42	0
027	Magnesium (%)	0675	0.3800	0.0000	0.3805	0.0207	0.0059	73	-0.03	027.43	0
027	Magnesium (%)	0407	0.3803	0.0050	0.3805	0.0207	0.0059	73	-0.01	027.43	0
027	Magnesium (%)	0572	0.3810	0.0080	0.3805	0.0207	0.0059	73	0.02	027.53	0
027	Magnesium (%)	0357	0.3817	0.0010	0.3805	0.0207	0.0059	73	0.06	027.42	0
027	Magnesium (%)	0948	0.3845	0.0190	0.3805	0.0207	0.0059	73	0.19	027.33	0
027	Magnesium (%)	0186	0.3847	0.0077	0.3805	0.0207	0.0059	73	0.20	027.42	0
027	Magnesium (%)	0910	0.3850	0.0100	0.3805	0.0207	0.0059	73	0.22	027.41	0
027	Magnesium (%)	0510	0.3850	0.0100	0.3805	0.0207	0.0059	73	0.22	027.43	0
027	Magnesium (%)	2089	0.3850	0.0100	0.3805	0.0207	0.0059	73	0.22	027.43	0
027	Magnesium (%)	0142	0.3851	0.0004	0.3805	0.0207	0.0059	73	0.22	027.31	0
027	Magnesium (%)	0870	0.3868	0.0070	0.3805	0.0207	0.0059	73	0.30	027.42	0
027	Magnesium (%)	0968	0.3875	0.0030	0.3805	0.0207	0.0059	73	0.34	027.43	0
027	Magnesium (%)	0560	0.3890	0.0085	0.3805	0.0207	0.0059	73	0.41	027.42	0
027	Magnesium (%)	0098	0.3900	0.0000	0.3805	0.0207	0.0059	73	0.46	027.41	0
027	Magnesium (%)	0098	0.3900	0.0000	0.3805	0.0207	0.0059	73	0.46	027.43	0
027	Magnesium (%)	0590	0.3900	0.0000	0.3805	0.0207	0.0059	73	0.46	027.99	0
027	Magnesium (%)	0017	0.3905	0.0010	0.3805	0.0207	0.0059	73	0.48	027.43	0
027	Magnesium (%)	2146	0.3945	0.0030	0.3805	0.0207	0.0059	73	0.68	027.51	0
027	Magnesium (%)	0598	0.3950	0.0004	0.3805	0.0207	0.0059	73	0.70	027.41	0
027	Magnesium (%)	0037	0.3950	0.0000	0.3805	0.0207	0.0059	73	0.70	027.42	0
027	Magnesium (%)	0964	0.3970	0.0007	0.3805	0.0207	0.0059	73	0.79	027.43	0
027	Magnesium (%)	0164	0.3970	0.0100	0.3805	0.0207	0.0059	73	0.80	027.41	0
027	Magnesium (%)	2196	0.4000	0.0000	0.3805	0.0207	0.0059	73	0.94	027.31	0
027	Magnesium (%)	0685	0.4000	0.0200	0.3805	0.0207	0.0059	73	0.94	027.42	0
027	Magnesium (%)	0553	0.4010	0.0020	0.3805	0.0207	0.0059	73	0.99	027.53	0
027	Magnesium (%)	2146	0.4030	0.0060	0.3805	0.0207	0.0059	73	1.09	027.31	0
027	Magnesium (%)	0520	0.4050	0.0100	0.3805	0.0207	0.0059	73	1.18	027.41	0
027	Magnesium (%)	0242	0.4050	0.0100	0.3805	0.0207	0.0059	73	1.18	027.99	0
027	Magnesium (%)	0171	0.4100	0.0000	0.3805	0.0207	0.0059	73	1.42	027.41	0
027	Magnesium (%)	0226	0.4100	0.0000	0.3805	0.0207	0.0059	73	1.42	027.41	0
027	Magnesium (%)	0693	0.4105	0.0170	0.3805	0.0207	0.0059	73	1.45	027.42	0
027	Magnesium (%)	0405	0.4195	0.0170	0.3805	0.0207	0.0059	73	1.88	027.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
027	Magnesium (%)	0682	0.4200	0.0000	0.3805	0.0207	0.0059	73	1.91	027.43	0
027	Magnesium (%)	0032	0.4218	0.0326	0.3805	0.0207	0.0059	73	1.99	027.43	0
027	Magnesium (%)	0100	0.4250	0.0100	0.3805	0.0207	0.0059	73	2.15	027.99	0
027	Magnesium (%)	0405	0.4305	0.0230	0.3805	0.0207	0.0059	73	2.42	027.42	0
027	Magnesium (%)	0042	0.3955	0.0530	0.3805	0.0207	0.0059	73	0.72	027.42	1
027	Magnesium (%)	0366	0.4000	0.0400	0.3805	0.0207	0.0059	73	0.94	027.42	1
027	Magnesium (%)	2246	0.1300	0.0000	0.3805	0.0207	0.0059	73	-12.11	027.31	2
027	Magnesium (%)	0199	< 0.5		0.3805	0.0207	0.0059	73		027.53	5
028	Manganese (mg / kg (ppm))	2248	56.70	0.4000	76.44	5.524	1.981	87	-3.57	028.31	0
028	Manganese (mg / kg (ppm))	0199	57.19	2.760	76.44	5.524	1.981	87	-3.48	028.53	0
028	Manganese (mg / kg (ppm))	0948	63.00	1.361	76.44	5.524	1.981	87	-2.43	028.33	0
028	Manganese (mg / kg (ppm))	0405	63.28	6.440	76.44	5.524	1.981	87	-2.38	028.42	0
028	Manganese (mg / kg (ppm))	0511	65.00	2.000	76.44	5.524	1.981	87	-2.07	028.41	0
028	Manganese (mg / kg (ppm))	0685	65.14	0.1500	76.44	5.524	1.981	87	-2.05	028.42	0
028	Manganese (mg / kg (ppm))	0175	65.50	1.000	76.44	5.524	1.981	87	-1.98	028.31	0
028	Manganese (mg / kg (ppm))	0169	67.50	2.200	76.44	5.524	1.981	87	-1.62	028.43	0
028	Manganese (mg / kg (ppm))	0171	67.75	0.3000	76.44	5.524	1.981	87	-1.57	028.41	0
028	Manganese (mg / kg (ppm))	0553	69.30	10.00	76.44	5.524	1.981	87	-1.29	028.53	0
028	Manganese (mg / kg (ppm))	0354	69.52	0.6700	76.44	5.524	1.981	87	-1.25	028.41	0
028	Manganese (mg / kg (ppm))	0939	70.02	0.1400	76.44	5.524	1.981	87	-1.16	028.53	0
028	Manganese (mg / kg (ppm))	0870	71.15	0.1000	76.44	5.524	1.981	87	-0.96	028.42	0
028	Manganese (mg / kg (ppm))	0610	71.50	1.000	76.44	5.524	1.981	87	-0.89	028.43	0
028	Manganese (mg / kg (ppm))	0263	71.70	0.6900	76.44	5.524	1.981	87	-0.86	028.42	0
028	Manganese (mg / kg (ppm))	0675	72.20	3.110	76.44	5.524	1.981	87	-0.77	028.43	0
028	Manganese (mg / kg (ppm))	0164	72.50	1.000	76.44	5.524	1.981	87	-0.71	028.41	0
028	Manganese (mg / kg (ppm))	2097	72.62	1.230	76.44	5.524	1.981	87	-0.69	028.31	0
028	Manganese (mg / kg (ppm))	0405	72.74	3.190	76.44	5.524	1.981	87	-0.67	028.41	0
028	Manganese (mg / kg (ppm))	0918	72.79	0.4100	76.44	5.524	1.981	87	-0.66	028.43	0
028	Manganese (mg / kg (ppm))	2196	72.90	0.0000	76.44	5.524	1.981	87	-0.64	028.31	0
028	Manganese (mg / kg (ppm))	0720	72.90	0.7600	76.44	5.524	1.981	87	-0.64	028.41	0
028	Manganese (mg / kg (ppm))	0017	73.00	2.466	76.44	5.524	1.981	87	-0.62	028.43	0
028	Manganese (mg / kg (ppm))	2129	73.20	0.6000	76.44	5.524	1.981	87	-0.59	028.99	0
028	Manganese (mg / kg (ppm))	0186	73.30	1.200	76.44	5.524	1.981	87	-0.57	028.52	0
028	Manganese (mg / kg (ppm))	0148	73.55	2.220	76.44	5.524	1.981	87	-0.52	028.41	0
028	Manganese (mg / kg (ppm))	0529	73.65	0.3000	76.44	5.524	1.981	87	-0.51	028.31	0
028	Manganese (mg / kg (ppm))	0074	74.00	2.000	76.44	5.524	1.981	87	-0.44	028.41	0
028	Manganese (mg / kg (ppm))	0278	74.00	4.000	76.44	5.524	1.981	87	-0.44	028.42	0
028	Manganese (mg / kg (ppm))	0353	74.00	2.000	76.44	5.524	1.981	87	-0.44	028.43	0
028	Manganese (mg / kg (ppm))	0049	74.28	1.640	76.44	5.524	1.981	87	-0.39	028.41	0
028	Manganese (mg / kg (ppm))	0536	74.33	0.2500	76.44	5.524	1.981	87	-0.38	028.31	0
028	Manganese (mg / kg (ppm))	0590	74.55	0.5000	76.44	5.524	1.981	87	-0.34	028.99	0
028	Manganese (mg / kg (ppm))	2012	74.61	0.3300	76.44	5.524	1.981	87	-0.33	028.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
028	Manganese (mg / kg (ppm))	0038	74.75	0.5000	76.44	5.524	1.981	87	-0.31	028.43	0
028	Manganese (mg / kg (ppm))	0939	75.02	0.1000	76.44	5.524	1.981	87	-0.26	028.31	0
028	Manganese (mg / kg (ppm))	0572	75.40	2.200	76.44	5.524	1.981	87	-0.19	028.53	0
028	Manganese (mg / kg (ppm))	0510	75.50	1.000	76.44	5.524	1.981	87	-0.17	028.43	0
028	Manganese (mg / kg (ppm))	0884	75.85	1.500	76.44	5.524	1.981	87	-0.11	028.31	0
028	Manganese (mg / kg (ppm))	0009	76.00	2.000	76.44	5.524	1.981	87	-0.08	028.42	0
028	Manganese (mg / kg (ppm))	0407	76.27	1.975	76.44	5.524	1.981	87	-0.03	028.41	0
028	Manganese (mg / kg (ppm))	0563	76.29	0.1302	76.44	5.524	1.981	87	-0.03	028.31	0
028	Manganese (mg / kg (ppm))	0520	76.50	1.000	76.44	5.524	1.981	87	0.01	028.41	0
028	Manganese (mg / kg (ppm))	0083	76.50	1.000	76.44	5.524	1.981	87	0.01	028.43	0
028	Manganese (mg / kg (ppm))	2192	76.51	0.1700	76.44	5.524	1.981	87	0.01	028.53	0
028	Manganese (mg / kg (ppm))	0226	76.80	1.210	76.44	5.524	1.981	87	0.06	028.41	0
028	Manganese (mg / kg (ppm))	0026	76.93	0.8900	76.44	5.524	1.981	87	0.09	028.42	0
028	Manganese (mg / kg (ppm))	0229	77.36	5.110	76.44	5.524	1.981	87	0.17	028.41	0
028	Manganese (mg / kg (ppm))	0265	77.45	6.900	76.44	5.524	1.981	87	0.18	028.42	0
028	Manganese (mg / kg (ppm))	0366	77.50	7.000	76.44	5.524	1.981	87	0.19	028.42	0
028	Manganese (mg / kg (ppm))	0242	77.50	3.000	76.44	5.524	1.981	87	0.19	028.99	0
028	Manganese (mg / kg (ppm))	0693	77.70	0.3240	76.44	5.524	1.981	87	0.23	028.42	0
028	Manganese (mg / kg (ppm))	0066	77.74	1.140	76.44	5.524	1.981	87	0.24	028.44	0
028	Manganese (mg / kg (ppm))	0001	77.80	2.660	76.44	5.524	1.981	87	0.25	028.31	0
028	Manganese (mg / kg (ppm))	0042	77.85	4.500	76.44	5.524	1.981	87	0.26	028.42	0
028	Manganese (mg / kg (ppm))	0186	78.50	1.000	76.44	5.524	1.981	87	0.37	028.42	0
028	Manganese (mg / kg (ppm))	0294	78.52	1.440	76.44	5.524	1.981	87	0.38	028.42	0
028	Manganese (mg / kg (ppm))	0505	78.96	2.960	76.44	5.524	1.981	87	0.46	028.43	0
028	Manganese (mg / kg (ppm))	0425	79.12	0.0500	76.44	5.524	1.981	87	0.49	028.43	0
028	Manganese (mg / kg (ppm))	0098	79.33	0.2200	76.44	5.524	1.981	87	0.52	028.41	0
028	Manganese (mg / kg (ppm))	0964	79.38	0.1017	76.44	5.524	1.981	87	0.53	028.43	0
028	Manganese (mg / kg (ppm))	0651	79.41	0.1900	76.44	5.524	1.981	87	0.54	028.31	0
028	Manganese (mg / kg (ppm))	0868	79.50	1.000	76.44	5.524	1.981	87	0.55	028.31	0
028	Manganese (mg / kg (ppm))	0560	79.50	2.400	76.44	5.524	1.981	87	0.55	028.42	0
028	Manganese (mg / kg (ppm))	0968	79.50	1.000	76.44	5.524	1.981	87	0.55	028.43	0
028	Manganese (mg / kg (ppm))	0682	79.78	0.0000	76.44	5.524	1.981	87	0.60	028.43	0
028	Manganese (mg / kg (ppm))	0598	79.79	0.0900	76.44	5.524	1.981	87	0.61	028.41	0
028	Manganese (mg / kg (ppm))	0045	79.80	2.800	76.44	5.524	1.981	87	0.61	028.42	0
028	Manganese (mg / kg (ppm))	0038	80.00	2.000	76.44	5.524	1.981	87	0.64	028.32	0
028	Manganese (mg / kg (ppm))	2089	80.14	1.520	76.44	5.524	1.981	87	0.67	028.43	0
028	Manganese (mg / kg (ppm))	0357	80.23	0.6830	76.44	5.524	1.981	87	0.69	028.42	0
028	Manganese (mg / kg (ppm))	0910	80.50	5.000	76.44	5.524	1.981	87	0.73	028.41	0
028	Manganese (mg / kg (ppm))	0098	81.77	1.790	76.44	5.524	1.981	87	0.96	028.43	0
028	Manganese (mg / kg (ppm))	0190	81.83	0.8000	76.44	5.524	1.981	87	0.98	028.42	0
028	Manganese (mg / kg (ppm))	0035	82.50	1.000	76.44	5.524	1.981	87	1.10	028.43	0
028	Manganese (mg / kg (ppm))	0100	82.50	1.000	76.44	5.524	1.981	87	1.10	028.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
028	Manganese (mg / kg (ppm))	0407	82.61	0.7956	76.44	5.524	1.981	87	1.12	028.43	0
028	Manganese (mg / kg (ppm))	0904	83.77	2.260	76.44	5.524	1.981	87	1.33	028.31	0
028	Manganese (mg / kg (ppm))	0027	84.73	3.497	76.44	5.524	1.981	87	1.50	028.43	0
028	Manganese (mg / kg (ppm))	0619	85.20	3.600	76.44	5.524	1.981	87	1.59	028.41	0
028	Manganese (mg / kg (ppm))	0032	85.21	10.55	76.44	5.524	1.981	87	1.59	028.43	0
028	Manganese (mg / kg (ppm))	0726	85.45	0.1900	76.44	5.524	1.981	87	1.63	028.42	0
028	Manganese (mg / kg (ppm))	0358	85.71	0.5700	76.44	5.524	1.981	87	1.68	028.41	0
028	Manganese (mg / kg (ppm))	0037	85.80	1.800	76.44	5.524	1.981	87	1.69	028.42	0
028	Manganese (mg / kg (ppm))	0202	86.20	3.035	76.44	5.524	1.981	87	1.77	028.43	0
028	Manganese (mg / kg (ppm))	0047	89.68	0.2900	76.44	5.524	1.981	87	2.40	028.52	0
028	Manganese (mg / kg (ppm))	0003	92.00	18.00	76.44	5.524	1.981	87	2.82	028.41	0
028	Manganese (mg / kg (ppm))	2128	53.60	63.45	76.44	5.524	1.981	87	-4.13	028.32	1
028	Manganese (mg / kg (ppm))	2188	141.0	1.310	76.44	5.524	1.981	87	11.69	028.31	2
031	Phosphorus (%)	2192	0.7350	0.0300	1.086	0.0478	0.0229	115	-7.34	031.01	0
031	Phosphorus (%)	2031	0.7534	0.0204	1.086	0.0478	0.0229	115	-6.96	031.01	0
031	Phosphorus (%)	0294	0.8800	0.0000	1.086	0.0478	0.0229	115	-4.31	031.42	0
031	Phosphorus (%)	0108	0.9000	0.0400	1.086	0.0478	0.0229	115	-3.89	031.01	0
031	Phosphorus (%)	0186	0.9656	0.0093	1.086	0.0478	0.0229	115	-2.51	031.52	0
031	Phosphorus (%)	0358	0.9700	0.0200	1.086	0.0478	0.0229	115	-2.42	031.41	0
031	Phosphorus (%)	0405	1.002	0.0130	1.086	0.0478	0.0229	115	-1.76	031.41	0
031	Phosphorus (%)	0154	1.007	0.0530	1.086	0.0478	0.0229	115	-1.64	031.52	0
031	Phosphorus (%)	2089	1.015	0.0100	1.086	0.0478	0.0229	115	-1.48	031.43	0
031	Phosphorus (%)	2259	1.020	0.0190	1.086	0.0478	0.0229	115	-1.38	031.41	0
031	Phosphorus (%)	0038	1.022	0.0380	1.086	0.0478	0.0229	115	-1.33	031.43	0
031	Phosphorus (%)	0693	1.023	0.1090	1.086	0.0478	0.0229	115	-1.32	031.42	0
031	Phosphorus (%)	0563	1.025	0.0128	1.086	0.0478	0.0229	115	-1.28	031.01	0
031	Phosphorus (%)	0511	1.025	0.0100	1.086	0.0478	0.0229	115	-1.27	031.01	0
031	Phosphorus (%)	2022	1.025	0.0100	1.086	0.0478	0.0229	115	-1.27	031.01	0
031	Phosphorus (%)	2146	1.025	0.0100	1.086	0.0478	0.0229	115	-1.27	031.51	0
031	Phosphorus (%)	0100	1.025	0.0100	1.086	0.0478	0.0229	115	-1.27	031.99	0
031	Phosphorus (%)	0354	1.030	0.0000	1.086	0.0478	0.0229	115	-1.16	031.41	0
031	Phosphorus (%)	0520	1.030	0.0000	1.086	0.0478	0.0229	115	-1.16	031.41	0
031	Phosphorus (%)	0278	1.030	0.0400	1.086	0.0478	0.0229	115	-1.16	031.42	0
031	Phosphorus (%)	0676	1.030	0.0000	1.086	0.0478	0.0229	115	-1.16	031.99	0
031	Phosphorus (%)	2259	1.031	0.0090	1.086	0.0478	0.0229	115	-1.15	031.42	0
031	Phosphorus (%)	0674	1.040	0.0000	1.086	0.0478	0.0229	115	-0.96	031.01	0
031	Phosphorus (%)	2146	1.040	0.0400	1.086	0.0478	0.0229	115	-0.96	031.01	0
031	Phosphorus (%)	0139	1.042	0.0050	1.086	0.0478	0.0229	115	-0.92	031.41	0
031	Phosphorus (%)	2248	1.050	0.0380	1.086	0.0478	0.0229	115	-0.75	031.01	0
031	Phosphorus (%)	0910	1.050	0.0400	1.086	0.0478	0.0229	115	-0.75	031.41	0
031	Phosphorus (%)	0685	1.055	0.0300	1.086	0.0478	0.0229	115	-0.64	031.41	0
031	Phosphorus (%)	0298	1.060	0.0200	1.086	0.0478	0.0229	115	-0.54	031.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	0366	1.060	0.1400	1.086	0.0478	0.0229	115	-0.54	031.42	0
031	Phosphorus (%)	0038	1.061	0.0530	1.086	0.0478	0.0229	115	-0.53	031.01	0
031	Phosphorus (%)	0918	1.065	0.0070	1.086	0.0478	0.0229	115	-0.44	031.43	0
031	Phosphorus (%)	0144	1.065	0.0100	1.086	0.0478	0.0229	115	-0.43	031.41	0
031	Phosphorus (%)	0720	1.065	0.0100	1.086	0.0478	0.0229	115	-0.43	031.41	0
031	Phosphorus (%)	0353	1.065	0.0100	1.086	0.0478	0.0229	115	-0.43	031.43	0
031	Phosphorus (%)	0142	1.069	0.0047	1.086	0.0478	0.0229	115	-0.35	031.01	0
031	Phosphorus (%)	2193	1.070	0.0000	1.086	0.0478	0.0229	115	-0.33	031.01	0
031	Phosphorus (%)	0089	1.070	0.0000	1.086	0.0478	0.0229	115	-0.33	031.43	0
031	Phosphorus (%)	0560	1.074	0.0122	1.086	0.0478	0.0229	115	-0.25	031.42	0
031	Phosphorus (%)	2018	1.075	0.0300	1.086	0.0478	0.0229	115	-0.22	031.01	0
031	Phosphorus (%)	0169	1.075	0.0900	1.086	0.0478	0.0229	115	-0.22	031.03	0
031	Phosphorus (%)	0148	1.075	0.0100	1.086	0.0478	0.0229	115	-0.22	031.41	0
031	Phosphorus (%)	0164	1.075	0.0100	1.086	0.0478	0.0229	115	-0.22	031.41	0
031	Phosphorus (%)	0610	1.075	0.0100	1.086	0.0478	0.0229	115	-0.22	031.43	0
031	Phosphorus (%)	0939	1.075	0.0500	1.086	0.0478	0.0229	115	-0.22	031.53	0
031	Phosphorus (%)	0874	1.077	0.0170	1.086	0.0478	0.0229	115	-0.19	031.01	0
031	Phosphorus (%)	0186	1.077	0.0041	1.086	0.0478	0.0229	115	-0.18	031.42	0
031	Phosphorus (%)	0265	1.080	0.0000	1.086	0.0478	0.0229	115	-0.12	031.42	0
031	Phosphorus (%)	0675	1.080	0.0000	1.086	0.0478	0.0229	115	-0.12	031.43	0
031	Phosphorus (%)	2128	1.081	0.0750	1.086	0.0478	0.0229	115	-0.11	031.01	0
031	Phosphorus (%)	0357	1.082	0.0129	1.086	0.0478	0.0229	115	-0.09	031.42	0
031	Phosphorus (%)	0017	1.083	0.0780	1.086	0.0478	0.0229	115	-0.05	031.43	0
031	Phosphorus (%)	0868	1.085	0.0100	1.086	0.0478	0.0229	115	-0.01	031.01	0
031	Phosphorus (%)	2246	1.085	0.0100	1.086	0.0478	0.0229	115	-0.01	031.01	0
031	Phosphorus (%)	0171	1.085	0.0100	1.086	0.0478	0.0229	115	-0.01	031.41	0
031	Phosphorus (%)	0045	1.085	0.0100	1.086	0.0478	0.0229	115	-0.01	031.42	0
031	Phosphorus (%)	0083	1.085	0.0100	1.086	0.0478	0.0229	115	-0.01	031.43	0
031	Phosphorus (%)	0229	1.087	0.0690	1.086	0.0478	0.0229	115	0.02	031.41	0
031	Phosphorus (%)	0964	1.090	0.0087	1.086	0.0478	0.0229	115	0.08	031.43	0
031	Phosphorus (%)	2196	1.090	0.0000	1.086	0.0478	0.0229	115	0.09	031.01	0
031	Phosphorus (%)	0651	1.091	0.0040	1.086	0.0478	0.0229	115	0.11	031.01	0
031	Phosphorus (%)	2062	1.094	0.0019	1.086	0.0478	0.0229	115	0.18	031.01	0
031	Phosphorus (%)	2129	1.095	0.0700	1.086	0.0478	0.0229	115	0.20	031.42	0
031	Phosphorus (%)	0590	1.095	0.0300	1.086	0.0478	0.0229	115	0.20	031.01	0
031	Phosphorus (%)	0049	1.095	0.0500	1.086	0.0478	0.0229	115	0.20	031.41	0
031	Phosphorus (%)	0042	1.095	0.0100	1.086	0.0478	0.0229	115	0.20	031.42	0
031	Phosphorus (%)	0687	1.100	0.1000	1.086	0.0478	0.0229	115	0.30	031.01	0
031	Phosphorus (%)	0897	1.100	0.0000	1.086	0.0478	0.0229	115	0.30	031.01	0
031	Phosphorus (%)	2251	1.100	0.0000	1.086	0.0478	0.0229	115	0.30	031.01	0
031	Phosphorus (%)	2012	1.100	0.0200	1.086	0.0478	0.0229	115	0.30	031.41	0
031	Phosphorus (%)	0035	1.100	0.0200	1.086	0.0478	0.0229	115	0.30	031.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	0263	1.103	0.0004	1.086	0.0478	0.0229	115	0.36	031.42	0
031	Phosphorus (%)	0881	1.104	0.0200	1.086	0.0478	0.0229	115	0.38	031.01	0
031	Phosphorus (%)	2190	1.105	0.0104	1.086	0.0478	0.0229	115	0.40	031.01	0
031	Phosphorus (%)	0904	1.105	0.0500	1.086	0.0478	0.0229	115	0.41	031.01	0
031	Phosphorus (%)	2188	1.105	0.0100	1.086	0.0478	0.0229	115	0.41	031.01	0
031	Phosphorus (%)	0553	1.105	0.0700	1.086	0.0478	0.0229	115	0.41	031.53	0
031	Phosphorus (%)	0425	1.106	0.0250	1.086	0.0478	0.0229	115	0.42	031.43	0
031	Phosphorus (%)	2097	1.107	0.0010	1.086	0.0478	0.0229	115	0.44	031.01	0
031	Phosphorus (%)	0626	1.107	0.0180	1.086	0.0478	0.0229	115	0.45	031.01	0
031	Phosphorus (%)	0885	1.108	0.0044	1.086	0.0478	0.0229	115	0.46	031.01	0
031	Phosphorus (%)	0870	1.109	0.0080	1.086	0.0478	0.0229	115	0.49	031.42	0
031	Phosphorus (%)	0683	1.110	0.0200	1.086	0.0478	0.0229	115	0.51	031.01	0
031	Phosphorus (%)	0689	1.110	0.0200	1.086	0.0478	0.0229	115	0.51	031.01	0
031	Phosphorus (%)	0939	1.110	0.0200	1.086	0.0478	0.0229	115	0.51	031.01	0
031	Phosphorus (%)	2198	1.110	0.0000	1.086	0.0478	0.0229	115	0.51	031.01	0
031	Phosphorus (%)	0682	1.110	0.0000	1.086	0.0478	0.0229	115	0.51	031.43	0
031	Phosphorus (%)	0407	1.111	0.0152	1.086	0.0478	0.0229	115	0.53	031.41	0
031	Phosphorus (%)	2006	1.115	0.0180	1.086	0.0478	0.0229	115	0.62	031.01	0
031	Phosphorus (%)	0036	1.117	0.0069	1.086	0.0478	0.0229	115	0.66	031.03	0
031	Phosphorus (%)	0884	1.120	0.0000	1.086	0.0478	0.0229	115	0.72	031.01	0
031	Phosphorus (%)	0968	1.122	0.0020	1.086	0.0478	0.0229	115	0.76	031.43	0
031	Phosphorus (%)	0001	1.123	0.0040	1.086	0.0478	0.0229	115	0.78	031.03	0
031	Phosphorus (%)	0098	1.124	0.0280	1.086	0.0478	0.0229	115	0.80	031.43	0
031	Phosphorus (%)	0098	1.128	0.0080	1.086	0.0478	0.0229	115	0.89	031.41	0
031	Phosphorus (%)	0536	1.130	0.0200	1.086	0.0478	0.0229	115	0.93	031.06	0
031	Phosphorus (%)	0199	1.130	0.0200	1.086	0.0478	0.0229	115	0.93	031.53	0
031	Phosphorus (%)	0202	1.131	0.0020	1.086	0.0478	0.0229	115	0.95	031.43	0
031	Phosphorus (%)	0407	1.133	0.0651	1.086	0.0478	0.0229	115	0.98	031.43	0
031	Phosphorus (%)	0035	1.135	0.0300	1.086	0.0478	0.0229	115	1.03	031.43	0
031	Phosphorus (%)	0948	1.140	0.0770	1.086	0.0478	0.0229	115	1.13	031.01	0
031	Phosphorus (%)	0242	1.140	0.0200	1.086	0.0478	0.0229	115	1.14	031.99	0
031	Phosphorus (%)	0726	1.145	0.0060	1.086	0.0478	0.0229	115	1.24	031.42	0
031	Phosphorus (%)	0510	1.145	0.0100	1.086	0.0478	0.0229	115	1.24	031.43	0
031	Phosphorus (%)	0622	1.146	0.0930	1.086	0.0478	0.0229	115	1.27	031.01	0
031	Phosphorus (%)	0190	1.155	0.0100	1.086	0.0478	0.0229	115	1.45	031.42	0
031	Phosphorus (%)	2109	1.162	0.0460	1.086	0.0478	0.0229	115	1.60	031.01	0
031	Phosphorus (%)	0226	1.170	0.0000	1.086	0.0478	0.0229	115	1.77	031.41	0
031	Phosphorus (%)	0572	1.170	0.0400	1.086	0.0478	0.0229	115	1.77	031.53	0
031	Phosphorus (%)	0598	1.183	0.0060	1.086	0.0478	0.0229	115	2.04	031.41	0
031	Phosphorus (%)	0619	1.185	0.0500	1.086	0.0478	0.0229	115	2.08	031.41	0
031	Phosphorus (%)	0037	1.258	0.0000	1.086	0.0478	0.0229	115	3.61	031.42	0
031	Phosphorus (%)	0405	1.273	0.0680	1.086	0.0478	0.0229	115	3.92	031.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	0852	1.300	0.0000	1.086	0.0478	0.0229	115	4.49	031.99	0
031	Phosphorus (%)	0032	1.331	0.0080	1.086	0.0478	0.0229	115	5.14	031.43	0
031	Phosphorus (%)	0003	1.475	0.7900	1.086	0.0478	0.0229	115	8.15	031.41	1
032	Potassium (%)	0169	1.160	0.0200	1.464	0.0771	0.0341	80	-3.94	032.32	0
032	Potassium (%)	0720	1.185	0.0100	1.464	0.0771	0.0341	80	-3.61	032.41	0
032	Potassium (%)	0089	1.220	0.0000	1.464	0.0771	0.0341	80	-3.16	032.43	0
032	Potassium (%)	0154	1.255	0.0516	1.464	0.0771	0.0341	80	-2.71	032.52	0
032	Potassium (%)	0964	1.263	0.0187	1.464	0.0771	0.0341	80	-2.60	032.43	0
032	Potassium (%)	0358	1.315	0.0500	1.464	0.0771	0.0341	80	-1.93	032.41	0
032	Potassium (%)	0425	1.330	0.0200	1.464	0.0771	0.0341	80	-1.73	032.43	0
032	Potassium (%)	0038	1.347	0.0010	1.464	0.0771	0.0341	80	-1.52	032.43	0
032	Potassium (%)	0675	1.360	0.0000	1.464	0.0771	0.0341	80	-1.34	032.43	0
032	Potassium (%)	0353	1.365	0.1300	1.464	0.0771	0.0341	80	-1.28	032.43	0
032	Potassium (%)	0186	1.375	0.0102	1.464	0.0771	0.0341	80	-1.15	032.52	0
032	Potassium (%)	2146	1.380	0.0800	1.464	0.0771	0.0341	80	-1.08	032.31	0
032	Potassium (%)	0968	1.383	0.0110	1.464	0.0771	0.0341	80	-1.05	032.43	0
032	Potassium (%)	2146	1.385	0.0100	1.464	0.0771	0.0341	80	-1.02	032.51	0
032	Potassium (%)	0265	1.390	0.0200	1.464	0.0771	0.0341	80	-0.95	032.42	0
032	Potassium (%)	0144	1.395	0.0100	1.464	0.0771	0.0341	80	-0.89	032.41	0
032	Potassium (%)	0405	1.402	0.0350	1.464	0.0771	0.0341	80	-0.80	032.41	0
032	Potassium (%)	0294	1.405	0.0100	1.464	0.0771	0.0341	80	-0.76	032.42	0
032	Potassium (%)	0148	1.410	0.0400	1.464	0.0771	0.0341	80	-0.69	032.41	0
032	Potassium (%)	0910	1.415	0.1500	1.464	0.0771	0.0341	80	-0.63	032.41	0
032	Potassium (%)	2129	1.415	0.0500	1.464	0.0771	0.0341	80	-0.63	032.42	0
032	Potassium (%)	0049	1.420	0.0400	1.464	0.0771	0.0341	80	-0.56	032.41	0
032	Potassium (%)	0045	1.420	0.0000	1.464	0.0771	0.0341	80	-0.56	032.42	0
032	Potassium (%)	0683	1.425	0.0300	1.464	0.0771	0.0341	80	-0.50	032.31	0
032	Potassium (%)	0904	1.425	0.1100	1.464	0.0771	0.0341	80	-0.50	032.31	0
032	Potassium (%)	0596	1.430	0.0000	1.464	0.0771	0.0341	80	-0.44	032.31	0
032	Potassium (%)	0354	1.430	0.0000	1.464	0.0771	0.0341	80	-0.44	032.41	0
032	Potassium (%)	0405	1.436	0.0870	1.464	0.0771	0.0341	80	-0.36	032.42	0
032	Potassium (%)	0278	1.440	0.0800	1.464	0.0771	0.0341	80	-0.31	032.42	0
032	Potassium (%)	0038	1.448	0.0134	1.464	0.0771	0.0341	80	-0.20	032.31	0
032	Potassium (%)	2012	1.450	0.0200	1.464	0.0771	0.0341	80	-0.18	032.41	0
032	Potassium (%)	0035	1.450	0.0400	1.464	0.0771	0.0341	80	-0.18	032.42	0
032	Potassium (%)	0139	1.451	0.0110	1.464	0.0771	0.0341	80	-0.17	032.41	0
032	Potassium (%)	0186	1.456	0.0259	1.464	0.0771	0.0341	80	-0.10	032.42	0
032	Potassium (%)	0868	1.460	0.0200	1.464	0.0771	0.0341	80	-0.05	032.31	0
032	Potassium (%)	0610	1.460	0.0000	1.464	0.0771	0.0341	80	-0.05	032.43	0
032	Potassium (%)	0870	1.462	0.0310	1.464	0.0771	0.0341	80	-0.03	032.42	0
032	Potassium (%)	0074	1.465	0.0100	1.464	0.0771	0.0341	80	0.02	032.41	0
032	Potassium (%)	0035	1.465	0.0100	1.464	0.0771	0.0341	80	0.02	032.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
032	Potassium (%)	0083	1.465	0.0300	1.464	0.0771	0.0341	80	0.02	032.43	0
032	Potassium (%)	0918	1.468	0.0090	1.464	0.0771	0.0341	80	0.05	032.43	0
032	Potassium (%)	0164	1.475	0.0300	1.464	0.0771	0.0341	80	0.15	032.41	0
032	Potassium (%)	0590	1.475	0.0700	1.464	0.0771	0.0341	80	0.15	032.99	0
032	Potassium (%)	2097	1.476	0.0427	1.464	0.0771	0.0341	80	0.16	032.31	0
032	Potassium (%)	0948	1.480	0.0180	1.464	0.0771	0.0341	80	0.21	032.31	0
032	Potassium (%)	0229	1.483	0.0240	1.464	0.0771	0.0341	80	0.25	032.41	0
032	Potassium (%)	0100	1.485	0.0500	1.464	0.0771	0.0341	80	0.28	032.99	0
032	Potassium (%)	0536	1.490	0.0000	1.464	0.0771	0.0341	80	0.34	032.31	0
032	Potassium (%)	0553	1.490	0.1200	1.464	0.0771	0.0341	80	0.34	032.53	0
032	Potassium (%)	0017	1.491	0.0600	1.464	0.0771	0.0341	80	0.36	032.43	0
032	Potassium (%)	0171	1.495	0.0100	1.464	0.0771	0.0341	80	0.41	032.41	0
032	Potassium (%)	0042	1.495	0.0300	1.464	0.0771	0.0341	80	0.41	032.42	0
032	Potassium (%)	0199	1.495	0.0300	1.464	0.0771	0.0341	80	0.41	032.53	0
032	Potassium (%)	0572	1.495	0.0500	1.464	0.0771	0.0341	80	0.41	032.53	0
032	Potassium (%)	0563	1.499	0.0547	1.464	0.0771	0.0341	80	0.46	032.31	0
032	Potassium (%)	0560	1.499	0.0152	1.464	0.0771	0.0341	80	0.46	032.42	0
032	Potassium (%)	0682	1.500	0.0000	1.464	0.0771	0.0341	80	0.47	032.43	0
032	Potassium (%)	0884	1.505	0.0100	1.464	0.0771	0.0341	80	0.54	032.02	0
032	Potassium (%)	0693	1.510	0.0120	1.464	0.0771	0.0341	80	0.60	032.42	0
032	Potassium (%)	0939	1.510	0.0800	1.464	0.0771	0.0341	80	0.60	032.53	0
032	Potassium (%)	0357	1.514	0.0062	1.464	0.0771	0.0341	80	0.66	032.42	0
032	Potassium (%)	0003	1.515	0.0700	1.464	0.0771	0.0341	80	0.67	032.41	0
032	Potassium (%)	2089	1.515	0.0100	1.464	0.0771	0.0341	80	0.67	032.43	0
032	Potassium (%)	0407	1.520	0.0296	1.464	0.0771	0.0341	80	0.73	032.41	0
032	Potassium (%)	0202	1.520	0.0340	1.464	0.0771	0.0341	80	0.73	032.43	0
032	Potassium (%)	0510	1.520	0.0000	1.464	0.0771	0.0341	80	0.73	032.43	0
032	Potassium (%)	2246	1.525	0.0300	1.464	0.0771	0.0341	80	0.80	032.32	0
032	Potassium (%)	0242	1.525	0.0100	1.464	0.0771	0.0341	80	0.80	032.99	0
032	Potassium (%)	0226	1.530	0.0600	1.464	0.0771	0.0341	80	0.86	032.41	0
032	Potassium (%)	0263	1.534	0.0152	1.464	0.0771	0.0341	80	0.92	032.42	0
032	Potassium (%)	0098	1.535	0.0300	1.464	0.0771	0.0341	80	0.93	032.43	0
032	Potassium (%)	0407	1.571	0.1342	1.464	0.0771	0.0341	80	1.39	032.43	0
032	Potassium (%)	0598	1.573	0.0120	1.464	0.0771	0.0341	80	1.42	032.41	0
032	Potassium (%)	0098	1.580	0.0000	1.464	0.0771	0.0341	80	1.51	032.41	0
032	Potassium (%)	0939	1.585	0.0100	1.464	0.0771	0.0341	80	1.57	032.31	0
032	Potassium (%)	0520	1.600	0.0400	1.464	0.0771	0.0341	80	1.77	032.41	0
032	Potassium (%)	0037	1.618	0.0150	1.464	0.0771	0.0341	80	2.00	032.42	0
032	Potassium (%)	0032	1.659	0.1140	1.464	0.0771	0.0341	80	2.53	032.43	0
032	Potassium (%)	0142	1.742	0.0758	1.464	0.0771	0.0341	80	3.61	032.31	0
032	Potassium (%)	0874	1.885	0.0300	1.464	0.0771	0.0341	80	5.46	032.31	0
032	Potassium (%)	0366	1.555	0.2100	1.464	0.0771	0.0341	80	1.19	032.42	1

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
032	Potassium (%)	0511	1.675	0.2300	1.464	0.0771	0.0341	80	2.74	032.41	1
033	Salt as chloride (%)	0144	0.0650	0.0100	0.1356	0.0318	0.0064	42	-2.22	033.03	0
033	Salt as chloride (%)	0897	0.0850	0.0100	0.1356	0.0318	0.0064	42	-1.59	033.00	0
033	Salt as chloride (%)	0083	0.0900	0.0000	0.1356	0.0318	0.0064	42	-1.43	033.01	0
033	Salt as chloride (%)	0885	0.0945	0.0017	0.1356	0.0318	0.0064	42	-1.29	033.00	0
033	Salt as chloride (%)	0425	0.0950	0.0100	0.1356	0.0318	0.0064	42	-1.28	033.01	0
033	Salt as chloride (%)	0045	0.1065	0.0110	0.1356	0.0318	0.0064	42	-0.91	033.00	0
033	Salt as chloride (%)	0868	0.1100	0.0000	0.1356	0.0318	0.0064	42	-0.80	033.05	0
033	Salt as chloride (%)	0650	0.1100	0.0000	0.1356	0.0318	0.0064	42	-0.80	033.99	0
033	Salt as chloride (%)	2062	0.1143	0.0031	0.1356	0.0318	0.0064	42	-0.67	033.00	0
033	Salt as chloride (%)	0366	0.1150	0.0100	0.1356	0.0318	0.0064	42	-0.65	033.00	0
033	Salt as chloride (%)	0148	0.1185	0.0070	0.1356	0.0318	0.0064	42	-0.54	033.01	0
033	Salt as chloride (%)	0539	0.1200	0.0000	0.1356	0.0318	0.0064	42	-0.49	033.00	0
033	Salt as chloride (%)	0354	0.1200	0.0000	0.1356	0.0318	0.0064	42	-0.49	033.01	0
033	Salt as chloride (%)	0682	0.1200	0.0000	0.1356	0.0318	0.0064	42	-0.49	033.99	0
033	Salt as chloride (%)	0407	0.1240	0.0000	0.1356	0.0318	0.0064	42	-0.36	033.01	0
033	Salt as chloride (%)	0590	0.1250	0.0100	0.1356	0.0318	0.0064	42	-0.33	033.01	0
033	Salt as chloride (%)	2109	0.1260	0.0020	0.1356	0.0318	0.0064	42	-0.30	033.01	0
033	Salt as chloride (%)	0884	0.1300	0.0000	0.1356	0.0318	0.0064	42	-0.18	033.00	0
033	Salt as chloride (%)	2022	0.1300	0.0200	0.1356	0.0318	0.0064	42	-0.18	033.00	0
033	Salt as chloride (%)	0164	0.1300	0.0000	0.1356	0.0318	0.0064	42	-0.18	033.01	0
033	Salt as chloride (%)	0242	0.1300	0.0000	0.1356	0.0318	0.0064	42	-0.18	033.01	0
033	Salt as chloride (%)	0948	0.1300	0.0000	0.1356	0.0318	0.0064	42	-0.18	033.01	0
033	Salt as chloride (%)	0693	0.1325	0.0090	0.1356	0.0318	0.0064	42	-0.10	033.00	0
033	Salt as chloride (%)	0881	0.1350	0.0060	0.1356	0.0318	0.0064	42	-0.02	033.00	0
033	Salt as chloride (%)	0683	0.1350	0.0100	0.1356	0.0318	0.0064	42	-0.02	033.01	0
033	Salt as chloride (%)	0278	0.1400	0.0000	0.1356	0.0318	0.0064	42	0.14	033.01	0
033	Salt as chloride (%)	0229	0.1410	0.0020	0.1356	0.0318	0.0064	42	0.17	033.01	0
033	Salt as chloride (%)	0353	0.1450	0.0100	0.1356	0.0318	0.0064	42	0.30	033.00	0
033	Salt as chloride (%)	0510	0.1480	0.0000	0.1356	0.0318	0.0064	42	0.39	033.01	0
033	Salt as chloride (%)	0689	0.1500	0.0000	0.1356	0.0318	0.0064	42	0.45	033.05	0
033	Salt as chloride (%)	0100	0.1500	0.0400	0.1356	0.0318	0.0064	42	0.45	033.01	0
033	Salt as chloride (%)	2190	0.1541	0.0132	0.1356	0.0318	0.0064	42	0.58	033.00	0
033	Salt as chloride (%)	0199	0.1550	0.0100	0.1356	0.0318	0.0064	42	0.61	033.01	0
033	Salt as chloride (%)	2076	0.1600	0.0000	0.1356	0.0318	0.0064	42	0.77	033.00	0
033	Salt as chloride (%)	0910	0.1650	0.0100	0.1356	0.0318	0.0064	42	0.92	033.99	0
033	Salt as chloride (%)	2073	0.1700	0.0000	0.1356	0.0318	0.0064	42	1.08	033.99	0
033	Salt as chloride (%)	2193	0.1850	0.0100	0.1356	0.0318	0.0064	42	1.55	033.00	0
033	Salt as chloride (%)	0874	0.1980	0.0000	0.1356	0.0318	0.0064	42	1.96	033.01	0
033	Salt as chloride (%)	0358	0.2350	0.0100	0.1356	0.0318	0.0064	42	3.12	033.99	0
033	Salt as chloride (%)	0017	0.2525	0.0090	0.1356	0.0318	0.0064	42	3.67	033.99	0
033	Salt as chloride (%)	0032	0.2770	0.0060	0.1356	0.0318	0.0064	42	4.44	033.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
033	Salt as chloride (%)	2043	0.2950	0.0300	0.1356	0.0318	0.0064	42	5.01	033.05	0
033	Salt as chloride (%)	0098	0.1550	0.0500	0.1356	0.0318	0.0064	42	0.61	033.01	1
033	Salt as chloride (%)	0226	0.1650	0.0700	0.1356	0.0318	0.0064	42	0.92	033.01	1
033	Salt as chloride (%)	0619	0.0000	0.0000	0.1356	0.0318	0.0064	42		033.01	4
033	Salt as chloride (%)	0265	< 0.12		0.1356	0.0318	0.0064	42		033.03	5
034	Selenium (mg / kg (ppm))	0171	0.2961	0.0042	0.3988	0.1028	0.0790	21	-1.00	034.04	0
034	Selenium (mg / kg (ppm))	0610	0.3100	0.0200	0.3988	0.1028	0.0790	21	-0.86	034.53	0
034	Selenium (mg / kg (ppm))	0003	0.3140	0.1060	0.3988	0.1028	0.0790	21	-0.83	034.53	0
034	Selenium (mg / kg (ppm))	0169	0.3150	0.0100	0.3988	0.1028	0.0790	21	-0.82	034.04	0
034	Selenium (mg / kg (ppm))	0164	0.3170	0.0080	0.3988	0.1028	0.0790	21	-0.80	034.53	0
034	Selenium (mg / kg (ppm))	0045	0.3245	0.0670	0.3988	0.1028	0.0790	21	-0.72	034.04	0
034	Selenium (mg / kg (ppm))	0619	0.3285	0.0230	0.3988	0.1028	0.0790	21	-0.68	034.41	0
034	Selenium (mg / kg (ppm))	0047	0.3350	0.0100	0.3988	0.1028	0.0790	21	-0.62	034.52	0
034	Selenium (mg / kg (ppm))	0148	0.3450	0.0300	0.3988	0.1028	0.0790	21	-0.52	034.41	0
034	Selenium (mg / kg (ppm))	0553	0.3695	0.0070	0.3988	0.1028	0.0790	21	-0.29	034.53	0
034	Selenium (mg / kg (ppm))	0038	0.3722	0.0082	0.3988	0.1028	0.0790	21	-0.26	034.01	0
034	Selenium (mg / kg (ppm))	0560	0.3740	0.0080	0.3988	0.1028	0.0790	21	-0.24	034.53	0
034	Selenium (mg / kg (ppm))	0910	0.3800	0.0600	0.3988	0.1028	0.0790	21	-0.18	034.52	0
034	Selenium (mg / kg (ppm))	0186	0.3961	0.0010	0.3988	0.1028	0.0790	21	-0.03	034.52	0
034	Selenium (mg / kg (ppm))	0939	0.4400	0.1400	0.3988	0.1028	0.0790	21	0.40	034.53	0
034	Selenium (mg / kg (ppm))	2012	0.4685	0.0210	0.3988	0.1028	0.0790	21	0.68	034.53	0
034	Selenium (mg / kg (ppm))	0918	0.5144	0.0262	0.3988	0.1028	0.0790	21	1.13	034.53	0
034	Selenium (mg / kg (ppm))	0032	0.5160	0.2300	0.3988	0.1028	0.0790	21	1.14	034.53	0
034	Selenium (mg / kg (ppm))	0199	0.6050	0.5300	0.3988	0.1028	0.0790	21	2.01	034.53	0
034	Selenium (mg / kg (ppm))	0407	1.310	0.3200	0.3988	0.1028	0.0790	21	8.87	034.43	0
034	Selenium (mg / kg (ppm))	0190	1.395	0.0300	0.3988	0.1028	0.0790	21	9.69	034.99	0
034	Selenium (mg / kg (ppm))	0407	1.685	0.7100	0.3988	0.1028	0.0790	21	12.52	034.41	1
035	Sodium (%)	2192	0.0800	0.0000	0.1071	0.0122	0.0044	88	-2.22	035.43	0
035	Sodium (%)	0169	0.0850	0.0100	0.1071	0.0122	0.0044	88	-1.81	035.32	0
035	Sodium (%)	0596	0.0900	0.0000	0.1071	0.0122	0.0044	88	-1.40	035.31	0
035	Sodium (%)	0144	0.0900	0.0000	0.1071	0.0122	0.0044	88	-1.40	035.41	0
035	Sodium (%)	0265	0.0900	0.0000	0.1071	0.0122	0.0044	88	-1.40	035.42	0
035	Sodium (%)	0294	0.0900	0.0000	0.1071	0.0122	0.0044	88	-1.40	035.42	0
035	Sodium (%)	0505	0.0900	0.0200	0.1071	0.0122	0.0044	88	-1.40	035.43	0
035	Sodium (%)	0964	0.0927	0.0002	0.1071	0.0122	0.0044	88	-1.18	035.43	0
035	Sodium (%)	0108	0.0950	0.0100	0.1071	0.0122	0.0044	88	-0.99	035.05	0
035	Sodium (%)	0175	0.0950	0.0100	0.1071	0.0122	0.0044	88	-0.99	035.31	0
035	Sodium (%)	0148	0.0950	0.0100	0.1071	0.0122	0.0044	88	-0.99	035.41	0
035	Sodium (%)	0358	0.0950	0.0100	0.1071	0.0122	0.0044	88	-0.99	035.41	0
035	Sodium (%)	0918	0.0960	0.0000	0.1071	0.0122	0.0044	88	-0.91	035.43	0
035	Sodium (%)	0553	0.0970	0.0060	0.1071	0.0122	0.0044	88	-0.83	035.53	0
035	Sodium (%)	0366	0.0975	0.0150	0.1071	0.0122	0.0044	88	-0.79	035.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
035	Sodium (%)	0263	0.0995	0.0010	0.1071	0.0122	0.0044	88	-0.62	035.42	0
035	Sodium (%)	0017	0.0995	0.0030	0.1071	0.0122	0.0044	88	-0.62	035.43	0
035	Sodium (%)	0939	0.1000	0.0000	0.1071	0.0122	0.0044	88	-0.58	035.31	0
035	Sodium (%)	2246	0.1000	0.0000	0.1071	0.0122	0.0044	88	-0.58	035.31	0
035	Sodium (%)	0049	0.1000	0.0000	0.1071	0.0122	0.0044	88	-0.58	035.41	0
035	Sodium (%)	0164	0.1000	0.0000	0.1071	0.0122	0.0044	88	-0.58	035.41	0
035	Sodium (%)	0298	0.1000	0.0200	0.1071	0.0122	0.0044	88	-0.58	035.41	0
035	Sodium (%)	0035	0.1000	0.0000	0.1071	0.0122	0.0044	88	-0.58	035.42	0
035	Sodium (%)	0045	0.1000	0.0000	0.1071	0.0122	0.0044	88	-0.58	035.42	0
035	Sodium (%)	0035	0.1000	0.0000	0.1071	0.0122	0.0044	88	-0.58	035.43	0
035	Sodium (%)	0098	0.1000	0.0000	0.1071	0.0122	0.0044	88	-0.58	035.43	0
035	Sodium (%)	0610	0.1000	0.0000	0.1071	0.0122	0.0044	88	-0.58	035.43	0
035	Sodium (%)	0682	0.1000	0.0000	0.1071	0.0122	0.0044	88	-0.58	035.43	0
035	Sodium (%)	0242	0.1000	0.0000	0.1071	0.0122	0.0044	88	-0.58	035.99	0
035	Sodium (%)	0154	0.1003	0.0043	0.1071	0.0122	0.0044	88	-0.56	035.52	0
035	Sodium (%)	0563	0.1012	0.0026	0.1071	0.0122	0.0044	88	-0.49	035.31	0
035	Sodium (%)	2146	0.1015	0.0010	0.1071	0.0122	0.0044	88	-0.46	035.51	0
035	Sodium (%)	0870	0.1019	0.0015	0.1071	0.0122	0.0044	88	-0.43	035.42	0
035	Sodium (%)	0510	0.1020	0.0000	0.1071	0.0122	0.0044	88	-0.42	035.43	0
035	Sodium (%)	0229	0.1025	0.0050	0.1071	0.0122	0.0044	88	-0.38	035.41	0
035	Sodium (%)	0560	0.1028	0.0005	0.1071	0.0122	0.0044	88	-0.36	035.42	0
035	Sodium (%)	0693	0.1035	0.0030	0.1071	0.0122	0.0044	88	-0.30	035.42	0
035	Sodium (%)	0038	0.1035	0.0090	0.1071	0.0122	0.0044	88	-0.30	035.43	0
035	Sodium (%)	0948	0.1040	0.0040	0.1071	0.0122	0.0044	88	-0.26	035.31	0
035	Sodium (%)	0405	0.1040	0.0040	0.1071	0.0122	0.0044	88	-0.26	035.41	0
035	Sodium (%)	2018	0.1050	0.0100	0.1071	0.0122	0.0044	88	-0.18	035.31	0
035	Sodium (%)	0098	0.1050	0.0100	0.1071	0.0122	0.0044	88	-0.18	035.41	0
035	Sodium (%)	0520	0.1050	0.0100	0.1071	0.0122	0.0044	88	-0.18	035.41	0
035	Sodium (%)	2012	0.1050	0.0100	0.1071	0.0122	0.0044	88	-0.18	035.41	0
035	Sodium (%)	0353	0.1050	0.0100	0.1071	0.0122	0.0044	88	-0.18	035.43	0
035	Sodium (%)	0100	0.1050	0.0100	0.1071	0.0122	0.0044	88	-0.18	035.99	0
035	Sodium (%)	0407	0.1051	0.0004	0.1071	0.0122	0.0044	88	-0.17	035.41	0
035	Sodium (%)	0598	0.1054	0.0003	0.1071	0.0122	0.0044	88	-0.15	035.41	0
035	Sodium (%)	0405	0.1065	0.0030	0.1071	0.0122	0.0044	88	-0.05	035.42	0
035	Sodium (%)	0142	0.1072	0.0081	0.1071	0.0122	0.0044	88	0.00	035.31	0
035	Sodium (%)	0042	0.1075	0.0090	0.1071	0.0122	0.0044	88	0.03	035.42	0
035	Sodium (%)	0186	0.1084	0.0038	0.1071	0.0122	0.0044	88	0.10	035.52	0
035	Sodium (%)	2146	0.1085	0.0010	0.1071	0.0122	0.0044	88	0.11	035.31	0
035	Sodium (%)	0619	0.1090	0.0040	0.1071	0.0122	0.0044	88	0.15	035.41	0
035	Sodium (%)	0202	0.1095	0.0030	0.1071	0.0122	0.0044	88	0.19	035.43	0
035	Sodium (%)	0884	0.1100	0.0000	0.1071	0.0122	0.0044	88	0.23	035.02	0
035	Sodium (%)	2248	0.1100	0.0000	0.1071	0.0122	0.0044	88	0.23	035.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
035	Sodium (%)	0171	0.1100	0.0000	0.1071	0.0122	0.0044	88	0.23	035.41	0
035	Sodium (%)	0226	0.1100	0.0000	0.1071	0.0122	0.0044	88	0.23	035.41	0
035	Sodium (%)	0354	0.1100	0.0000	0.1071	0.0122	0.0044	88	0.23	035.41	0
035	Sodium (%)	0910	0.1100	0.0200	0.1071	0.0122	0.0044	88	0.23	035.41	0
035	Sodium (%)	0083	0.1100	0.0000	0.1071	0.0122	0.0044	88	0.23	035.43	0
035	Sodium (%)	0939	0.1100	0.0000	0.1071	0.0122	0.0044	88	0.23	035.53	0
035	Sodium (%)	0590	0.1100	0.0000	0.1071	0.0122	0.0044	88	0.23	035.99	0
035	Sodium (%)	0881	0.1105	0.0030	0.1071	0.0122	0.0044	88	0.27	035.43	0
035	Sodium (%)	0407	0.1129	0.0012	0.1071	0.0122	0.0044	88	0.47	035.43	0
035	Sodium (%)	0529	0.1145	0.0030	0.1071	0.0122	0.0044	88	0.60	035.31	0
035	Sodium (%)	0904	0.1150	0.0100	0.1071	0.0122	0.0044	88	0.64	035.31	0
035	Sodium (%)	0720	0.1150	0.0100	0.1071	0.0122	0.0044	88	0.64	035.41	0
035	Sodium (%)	2251	0.1200	0.0000	0.1071	0.0122	0.0044	88	1.05	035.05	0
035	Sodium (%)	2196	0.1200	0.0000	0.1071	0.0122	0.0044	88	1.05	035.31	0
035	Sodium (%)	0278	0.1200	0.0200	0.1071	0.0122	0.0044	88	1.05	035.42	0
035	Sodium (%)	2089	0.1200	0.0000	0.1071	0.0122	0.0044	88	1.05	035.43	0
035	Sodium (%)	2097	0.1213	0.0012	0.1071	0.0122	0.0044	88	1.16	035.31	0
035	Sodium (%)	0037	0.1250	0.0040	0.1071	0.0122	0.0044	88	1.46	035.05	0
035	Sodium (%)	0425	0.1250	0.0100	0.1071	0.0122	0.0044	88	1.46	035.43	0
035	Sodium (%)	0186	0.1262	0.0052	0.1071	0.0122	0.0044	88	1.56	035.42	0
035	Sodium (%)	0622	0.1289	0.0039	0.1071	0.0122	0.0044	88	1.77	035.31	0
035	Sodium (%)	0868	0.1300	0.0000	0.1071	0.0122	0.0044	88	1.87	035.31	0
035	Sodium (%)	0675	0.1300	0.0200	0.1071	0.0122	0.0044	88	1.87	035.43	0
035	Sodium (%)	0868	0.1365	0.0010	0.1071	0.0122	0.0044	88	2.40	035.01	0
035	Sodium (%)	0874	0.1400	0.0000	0.1071	0.0122	0.0044	88	2.68	035.31	0
035	Sodium (%)	0139	0.1495	0.0010	0.1071	0.0122	0.0044	88	3.46	035.41	0
035	Sodium (%)	0038	0.1925	0.0025	0.1071	0.0122	0.0044	88	6.97	035.31	0
035	Sodium (%)	0089	0.2000	0.0000	0.1071	0.0122	0.0044	88	7.58	035.43	0
035	Sodium (%)	0683	0.2350	0.0100	0.1071	0.0122	0.0044	88	10.44	035.31	0
035	Sodium (%)	0572	0.9755	0.0190	0.1071	0.0122	0.0044	88	70.93	035.53	0
035	Sodium (%)	2129	0.9865	0.0030	0.1071	0.0122	0.0044	88	71.83	035.99	0
035	Sodium (%)	0536	4.035	7.930	0.1071	0.0122	0.0044	88	320.85	035.31	2
035	Sodium (%)	0199	< 0.1		0.1071	0.0122	0.0044	88		035.53	5
036	Sulfur (%)	0037	0.1615	0.0010	0.3654	0.0275	0.0085	38	-7.41	036.42	0
036	Sulfur (%)	0229	0.2750	0.0100	0.3654	0.0275	0.0085	38	-3.29	036.42	0
036	Sulfur (%)	0294	0.3150	0.0100	0.3654	0.0275	0.0085	38	-1.83	036.42	0
036	Sulfur (%)	0353	0.3250	0.0100	0.3654	0.0275	0.0085	38	-1.47	036.43	0
036	Sulfur (%)	0405	0.3290	0.0280	0.3654	0.0275	0.0085	38	-1.32	036.42	0
036	Sulfur (%)	0366	0.3300	0.0400	0.3654	0.0275	0.0085	38	-1.29	036.42	0
036	Sulfur (%)	0278	0.3300	0.0200	0.3654	0.0275	0.0085	38	-1.29	036.42	0
036	Sulfur (%)	0870	0.3437	0.0078	0.3654	0.0275	0.0085	38	-0.79	036.42	0
036	Sulfur (%)	0918	0.3530	0.0020	0.3654	0.0275	0.0085	38	-0.45	036.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
036	Sulfur (%)	0910	0.3547	0.0027	0.3654	0.0275	0.0085	38	-0.39	036.52	0
036	Sulfur (%)	0038	0.3565	0.0010	0.3654	0.0275	0.0085	38	-0.32	036.43	0
036	Sulfur (%)	0186	0.3566	0.0034	0.3654	0.0275	0.0085	38	-0.32	036.52	0
036	Sulfur (%)	0045	0.3590	0.0020	0.3654	0.0275	0.0085	38	-0.23	036.42	0
036	Sulfur (%)	0357	0.3598	0.0027	0.3654	0.0275	0.0085	38	-0.21	036.42	0
036	Sulfur (%)	0035	0.3600	0.0000	0.3654	0.0275	0.0085	38	-0.20	036.43	0
036	Sulfur (%)	0083	0.3600	0.0000	0.3654	0.0275	0.0085	38	-0.20	036.43	0
036	Sulfur (%)	0242	0.3600	0.0000	0.3654	0.0275	0.0085	38	-0.20	036.99	0
036	Sulfur (%)	0265	0.3650	0.0100	0.3654	0.0275	0.0085	38	-0.02	036.42	0
036	Sulfur (%)	0169	0.3650	0.0100	0.3654	0.0275	0.0085	38	-0.02	036.43	0
036	Sulfur (%)	2129	0.3660	0.0160	0.3654	0.0275	0.0085	38	0.02	036.42	0
036	Sulfur (%)	0186	0.3668	0.0079	0.3654	0.0275	0.0085	38	0.05	036.42	0
036	Sulfur (%)	0693	0.3670	0.0000	0.3654	0.0275	0.0085	38	0.06	036.42	0
036	Sulfur (%)	0164	0.3700	0.0000	0.3654	0.0275	0.0085	38	0.17	036.42	0
036	Sulfur (%)	0407	0.3718	0.0010	0.3654	0.0275	0.0085	38	0.23	036.43	0
036	Sulfur (%)	0560	0.3746	0.0117	0.3654	0.0275	0.0085	38	0.33	036.42	0
036	Sulfur (%)	0708	0.3775	0.0030	0.3654	0.0275	0.0085	38	0.44	036.42	0
036	Sulfur (%)	0553	0.3805	0.0310	0.3654	0.0275	0.0085	38	0.55	036.53	0
036	Sulfur (%)	0229	0.3815	0.0130	0.3654	0.0275	0.0085	38	0.58	036.04	0
036	Sulfur (%)	0148	0.3890	0.0020	0.3654	0.0275	0.0085	38	0.86	036.04	0
036	Sulfur (%)	0202	0.3895	0.0050	0.3654	0.0275	0.0085	38	0.87	036.43	0
036	Sulfur (%)	0098	0.3897	0.0127	0.3654	0.0275	0.0085	38	0.88	036.43	0
036	Sulfur (%)	0726	0.3898	0.0001	0.3654	0.0275	0.0085	38	0.88	036.42	0
036	Sulfur (%)	0035	0.3900	0.0200	0.3654	0.0275	0.0085	38	0.89	036.42	0
036	Sulfur (%)	0510	0.3900	0.0000	0.3654	0.0275	0.0085	38	0.89	036.43	0
036	Sulfur (%)	0190	0.3950	0.0100	0.3654	0.0275	0.0085	38	1.07	036.42	0
036	Sulfur (%)	2089	0.4050	0.0100	0.3654	0.0275	0.0085	38	1.44	036.43	0
036	Sulfur (%)	0171	0.4200	0.0000	0.3654	0.0275	0.0085	38	1.98	036.42	0
036	Sulfur (%)	0226	0.4400	0.0200	0.3654	0.0275	0.0085	38	2.71	036.04	0
036	Sulfur (%)	0032	0.3860	0.0528	0.3654	0.0275	0.0085	38	0.75	036.43	1
037	Zinc (mg / kg (ppm))	0278	69.60	4.400	101.2	8.208	4.250	97	-3.85	037.42	0
037	Zinc (mg / kg (ppm))	0610	81.50	1.000	101.2	8.208	4.250	97	-2.40	037.43	0
037	Zinc (mg / kg (ppm))	0870	86.79	3.170	101.2	8.208	4.250	97	-1.76	037.42	0
037	Zinc (mg / kg (ppm))	0294	88.00	2.400	101.2	8.208	4.250	97	-1.61	037.42	0
037	Zinc (mg / kg (ppm))	0918	89.61	0.0700	101.2	8.208	4.250	97	-1.42	037.43	0
037	Zinc (mg / kg (ppm))	0511	91.00	4.000	101.2	8.208	4.250	97	-1.25	037.41	0
037	Zinc (mg / kg (ppm))	0017	91.88	5.107	101.2	8.208	4.250	97	-1.14	037.43	0
037	Zinc (mg / kg (ppm))	0529	91.95	2.900	101.2	8.208	4.250	97	-1.13	037.31	0
037	Zinc (mg / kg (ppm))	0366	92.00	6.000	101.2	8.208	4.250	97	-1.12	037.42	0
037	Zinc (mg / kg (ppm))	0939	92.02	0.8900	101.2	8.208	4.250	97	-1.12	037.53	0
037	Zinc (mg / kg (ppm))	0358	92.87	4.000	101.2	8.208	4.250	97	-1.02	037.41	0
037	Zinc (mg / kg (ppm))	0164	93.00	6.000	101.2	8.208	4.250	97	-1.00	037.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
037	Zinc (mg / kg (ppm))	0242	93.00	0.0000	101.2	8.208	4.250	97	-1.00	037.99	0
037	Zinc (mg / kg (ppm))	0265	93.30	6.000	101.2	8.208	4.250	97	-0.97	037.42	0
037	Zinc (mg / kg (ppm))	2129	93.40	0.2000	101.2	8.208	4.250	97	-0.95	037.99	0
037	Zinc (mg / kg (ppm))	2062	93.92	1.000	101.2	8.208	4.250	97	-0.89	037.31	0
037	Zinc (mg / kg (ppm))	0186	94.00	4.000	101.2	8.208	4.250	97	-0.88	037.42	0
037	Zinc (mg / kg (ppm))	0848	94.34	0.4500	101.2	8.208	4.250	97	-0.84	037.41	0
037	Zinc (mg / kg (ppm))	0354	94.37	0.0300	101.2	8.208	4.250	97	-0.84	037.41	0
037	Zinc (mg / kg (ppm))	0884	94.50	1.400	101.2	8.208	4.250	97	-0.82	037.31	0
037	Zinc (mg / kg (ppm))	0939	94.66	0.0000	101.2	8.208	4.250	97	-0.80	037.31	0
037	Zinc (mg / kg (ppm))	0263	94.89	0.2600	101.2	8.208	4.250	97	-0.77	037.42	0
037	Zinc (mg / kg (ppm))	2097	95.00	0.0000	101.2	8.208	4.250	97	-0.76	037.31	0
037	Zinc (mg / kg (ppm))	2196	95.10	0.0000	101.2	8.208	4.250	97	-0.75	037.31	0
037	Zinc (mg / kg (ppm))	0226	95.23	8.220	101.2	8.208	4.250	97	-0.73	037.41	0
037	Zinc (mg / kg (ppm))	0169	95.25	6.500	101.2	8.208	4.250	97	-0.73	037.43	0
037	Zinc (mg / kg (ppm))	0675	95.26	10.95	101.2	8.208	4.250	97	-0.73	037.43	0
037	Zinc (mg / kg (ppm))	0083	95.50	1.000	101.2	8.208	4.250	97	-0.70	037.43	0
037	Zinc (mg / kg (ppm))	0720	95.78	1.040	101.2	8.208	4.250	97	-0.66	037.41	0
037	Zinc (mg / kg (ppm))	0904	96.45	3.100	101.2	8.208	4.250	97	-0.58	037.31	0
037	Zinc (mg / kg (ppm))	0009	97.00	8.000	101.2	8.208	4.250	97	-0.52	037.42	0
037	Zinc (mg / kg (ppm))	0049	97.14	2.870	101.2	8.208	4.250	97	-0.50	037.41	0
037	Zinc (mg / kg (ppm))	0682	97.34	0.0000	101.2	8.208	4.250	97	-0.47	037.43	0
037	Zinc (mg / kg (ppm))	0948	97.45	0.8160	101.2	8.208	4.250	97	-0.46	037.33	0
037	Zinc (mg / kg (ppm))	2248	97.69	2.410	101.2	8.208	4.250	97	-0.43	037.31	0
037	Zinc (mg / kg (ppm))	2012	97.69	2.760	101.2	8.208	4.250	97	-0.43	037.41	0
037	Zinc (mg / kg (ppm))	0066	97.73	2.120	101.2	8.208	4.250	97	-0.43	037.44	0
037	Zinc (mg / kg (ppm))	0014	98.10	2.800	101.2	8.208	4.250	97	-0.38	037.32	0
037	Zinc (mg / kg (ppm))	2022	98.47	0.4490	101.2	8.208	4.250	97	-0.34	037.31	0
037	Zinc (mg / kg (ppm))	0868	98.50	1.000	101.2	8.208	4.250	97	-0.33	037.31	0
037	Zinc (mg / kg (ppm))	2089	98.71	1.590	101.2	8.208	4.250	97	-0.31	037.43	0
037	Zinc (mg / kg (ppm))	0536	99.38	1.950	101.2	8.208	4.250	97	-0.23	037.31	0
037	Zinc (mg / kg (ppm))	2146	99.54	2.670	101.2	8.208	4.250	97	-0.21	037.51	0
037	Zinc (mg / kg (ppm))	0229	99.65	13.71	101.2	8.208	4.250	97	-0.19	037.41	0
037	Zinc (mg / kg (ppm))	0010	100.0	0.0000	101.2	8.208	4.250	97	-0.15	037.33	0
037	Zinc (mg / kg (ppm))	0553	100.3	5.400	101.2	8.208	4.250	97	-0.11	037.53	0
037	Zinc (mg / kg (ppm))	0074	100.5	15.00	101.2	8.208	4.250	97	-0.09	037.41	0
037	Zinc (mg / kg (ppm))	0968	100.5	1.000	101.2	8.208	4.250	97	-0.09	037.43	0
037	Zinc (mg / kg (ppm))	0148	100.6	2.040	101.2	8.208	4.250	97	-0.08	037.41	0
037	Zinc (mg / kg (ppm))	0045	100.8	6.400	101.2	8.208	4.250	97	-0.05	037.42	0
037	Zinc (mg / kg (ppm))	0505	100.9	19.33	101.2	8.208	4.250	97	-0.04	037.43	0
037	Zinc (mg / kg (ppm))	0693	101.1	2.780	101.2	8.208	4.250	97	-0.02	037.42	0
037	Zinc (mg / kg (ppm))	0199	101.1	2.170	101.2	8.208	4.250	97	-0.01	037.53	0
037	Zinc (mg / kg (ppm))	0964	101.6	0.9440	101.2	8.208	4.250	97	0.05	037.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
037	Zinc (mg / kg (ppm))	0560	101.9	10.10	101.2	8.208	4.250	97	0.08	037.42	0
037	Zinc (mg / kg (ppm))	0619	102.0	4.000	101.2	8.208	4.250	97	0.09	037.41	0
037	Zinc (mg / kg (ppm))	2192	102.2	7.640	101.2	8.208	4.250	97	0.11	037.43	0
037	Zinc (mg / kg (ppm))	0598	102.6	0.4000	101.2	8.208	4.250	97	0.17	037.41	0
037	Zinc (mg / kg (ppm))	0001	102.7	2.400	101.2	8.208	4.250	97	0.18	037.31	0
037	Zinc (mg / kg (ppm))	0590	102.9	8.300	101.2	8.208	4.250	97	0.20	037.99	0
037	Zinc (mg / kg (ppm))	0572	103.0	2.000	101.2	8.208	4.250	97	0.22	037.53	0
037	Zinc (mg / kg (ppm))	0171	103.1	0.3000	101.2	8.208	4.250	97	0.22	037.41	0
037	Zinc (mg / kg (ppm))	2146	103.4	2.280	101.2	8.208	4.250	97	0.27	037.31	0
037	Zinc (mg / kg (ppm))	0425	103.5	1.050	101.2	8.208	4.250	97	0.28	037.43	0
037	Zinc (mg / kg (ppm))	0563	103.6	6.102	101.2	8.208	4.250	97	0.29	037.31	0
037	Zinc (mg / kg (ppm))	0874	104.0	4.000	101.2	8.208	4.250	97	0.34	037.31	0
037	Zinc (mg / kg (ppm))	0026	104.3	1.198	101.2	8.208	4.250	97	0.37	037.42	0
037	Zinc (mg / kg (ppm))	2246	104.5	1.000	101.2	8.208	4.250	97	0.40	037.33	0
037	Zinc (mg / kg (ppm))	0520	104.5	1.000	101.2	8.208	4.250	97	0.40	037.41	0
037	Zinc (mg / kg (ppm))	0038	105.9	12.30	101.2	8.208	4.250	97	0.56	037.32	0
037	Zinc (mg / kg (ppm))	0042	106.0	2.000	101.2	8.208	4.250	97	0.58	037.42	0
037	Zinc (mg / kg (ppm))	0510	106.0	0.0000	101.2	8.208	4.250	97	0.58	037.43	0
037	Zinc (mg / kg (ppm))	0038	106.1	9.900	101.2	8.208	4.250	97	0.59	037.43	0
037	Zinc (mg / kg (ppm))	0175	106.5	3.000	101.2	8.208	4.250	97	0.64	037.31	0
037	Zinc (mg / kg (ppm))	0202	106.8	7.634	101.2	8.208	4.250	97	0.68	037.43	0
037	Zinc (mg / kg (ppm))	0098	107.0	5.900	101.2	8.208	4.250	97	0.70	037.41	0
037	Zinc (mg / kg (ppm))	2178	107.5	25.00	101.2	8.208	4.250	97	0.76	037.33	0
037	Zinc (mg / kg (ppm))	0407	107.5	0.6631	101.2	8.208	4.250	97	0.77	037.41	0
037	Zinc (mg / kg (ppm))	0357	107.8	1.317	101.2	8.208	4.250	97	0.80	037.42	0
037	Zinc (mg / kg (ppm))	0190	109.4	3.420	101.2	8.208	4.250	97	1.00	037.42	0
037	Zinc (mg / kg (ppm))	0726	110.3	0.3000	101.2	8.208	4.250	97	1.10	037.42	0
037	Zinc (mg / kg (ppm))	0032	110.9	0.3000	101.2	8.208	4.250	97	1.17	037.43	0
037	Zinc (mg / kg (ppm))	0027	111.6	1.576	101.2	8.208	4.250	97	1.27	037.43	0
037	Zinc (mg / kg (ppm))	0004	111.8	11.58	101.2	8.208	4.250	97	1.29	037.41	0
037	Zinc (mg / kg (ppm))	0910	112.5	1.000	101.2	8.208	4.250	97	1.37	037.41	0
037	Zinc (mg / kg (ppm))	0035	113.0	2.000	101.2	8.208	4.250	97	1.43	037.43	0
037	Zinc (mg / kg (ppm))	0186	114.5	13.00	101.2	8.208	4.250	97	1.62	037.52	0
037	Zinc (mg / kg (ppm))	0035	115.0	2.000	101.2	8.208	4.250	97	1.68	037.42	0
037	Zinc (mg / kg (ppm))	0100	115.0	2.000	101.2	8.208	4.250	97	1.68	037.99	0
037	Zinc (mg / kg (ppm))	0098	115.3	4.100	101.2	8.208	4.250	97	1.71	037.43	0
037	Zinc (mg / kg (ppm))	0353	117.5	1.000	101.2	8.208	4.250	97	1.98	037.43	0
037	Zinc (mg / kg (ppm))	0037	118.3	2.200	101.2	8.208	4.250	97	2.08	037.42	0
037	Zinc (mg / kg (ppm))	0407	122.6	9.910	101.2	8.208	4.250	97	2.61	037.43	0
037	Zinc (mg / kg (ppm))	0405	124.7	4.070	101.2	8.208	4.250	97	2.87	037.41	0
037	Zinc (mg / kg (ppm))	0003	127.0	34.00	101.2	8.208	4.250	97	3.14	037.41	0
037	Zinc (mg / kg (ppm))	0047	154.6	9.700	101.2	8.208	4.250	97	6.50	037.52	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
037	Zinc (mg / kg (ppm))	0405	198.0	8.330	101.2	8.208	4.250	97	11.79	037.42	0
037	Zinc (mg / kg (ppm))	2128	39.65	78.44	101.2	8.208	4.250	97	-7.50	037.32	1
037	Zinc (mg / kg (ppm))	2188	368.4	105.4	101.2	8.208	4.250	97	32.55	037.31	2
038	Molybdenum (mg / kg (ppm))	0171	1.475	0.0100	1.771	0.2107	0.1161	17	-1.41	038.41	0
038	Molybdenum (mg / kg (ppm))	0186	1.490	0.1200	1.771	0.2107	0.1161	17	-1.34	038.52	0
038	Molybdenum (mg / kg (ppm))	0910	1.560	0.1000	1.771	0.2107	0.1161	17	-1.00	038.52	0
038	Molybdenum (mg / kg (ppm))	0278	1.620	0.1600	1.771	0.2107	0.1161	17	-0.72	038.42	0
038	Molybdenum (mg / kg (ppm))	0407	1.635	0.2621	1.771	0.2107	0.1161	17	-0.65	038.41	0
038	Molybdenum (mg / kg (ppm))	0964	1.679	0.0056	1.771	0.2107	0.1161	17	-0.44	038.43	0
038	Molybdenum (mg / kg (ppm))	0407	1.721	0.4000	1.771	0.2107	0.1161	17	-0.24	038.43	0
038	Molybdenum (mg / kg (ppm))	0164	1.775	0.0500	1.771	0.2107	0.1161	17	0.02	038.53	0
038	Molybdenum (mg / kg (ppm))	0083	1.800	0.0000	1.771	0.2107	0.1161	17	0.14	038.43	0
038	Molybdenum (mg / kg (ppm))	0169	1.800	0.0600	1.771	0.2107	0.1161	17	0.14	038.43	0
038	Molybdenum (mg / kg (ppm))	0510	1.800	0.0000	1.771	0.2107	0.1161	17	0.14	038.43	0
038	Molybdenum (mg / kg (ppm))	0693	1.811	0.0470	1.771	0.2107	0.1161	17	0.19	038.42	0
038	Molybdenum (mg / kg (ppm))	0918	1.829	0.0092	1.771	0.2107	0.1161	17	0.27	038.53	0
038	Molybdenum (mg / kg (ppm))	0553	1.970	0.1400	1.771	0.2107	0.1161	17	0.94	038.53	0
038	Molybdenum (mg / kg (ppm))	0226	1.975	0.0900	1.771	0.2107	0.1161	17	0.97	038.41	0
038	Molybdenum (mg / kg (ppm))	0353	2.173	0.0698	1.771	0.2107	0.1161	17	1.91	038.43	0
038	Molybdenum (mg / kg (ppm))	0045	2.465	0.4500	1.771	0.2107	0.1161	17	3.29	038.42	0
038	Molybdenum (mg / kg (ppm))	0038	4.972	1.876	1.771	0.2107	0.1161	17	15.19	038.43	1
040	Barium (mg / kg (ppm))	0918	14.18	0.2986				2		040.53	0
040	Barium (mg / kg (ppm))	0560	15.97	0.6700				2		040.42	0
041	Vanadium (mg / kg (ppm))	0553	1.230	0.1200				1		041.53	0
042	Chloride (%)	0948	0.0790	0.0000				1		042.00	0
102	Niacin (mg / kg (ppm))	0227	111.5	1.000				1		102.01	0
103	Pantothenic Acid (mg / kg (ppm))	0227	13.25	0.9000				1		103.01	0
104	Riboflavin (mg / kg (ppm))	0910	1.545	0.2100	3.004	1.234	0.2425	4		104.03	0
104	Riboflavin (mg / kg (ppm))	0010	2.450	0.1000	3.004	1.234	0.2425	4		104.03	0
104	Riboflavin (mg / kg (ppm))	0227	3.780	0.5800	3.004	1.234	0.2425	4		104.00	0
104	Riboflavin (mg / kg (ppm))	0939	4.240	0.0800	3.004	1.234	0.2425	4		104.03	0
105	Thiamine (mg / kg (ppm))	0010	4.300	0.2000	6.039	2.178	0.4625	4		105.00	0
105	Thiamine (mg / kg (ppm))	0910	4.580	0.2600	6.039	2.178	0.4625	4		105.00	0
105	Thiamine (mg / kg (ppm))	0227	6.230	0.8400	6.039	2.178	0.4625	4		105.01	0
105	Thiamine (mg / kg (ppm))	0939	9.045	0.5500	6.039	2.178	0.4625	4		105.00	0
106	Vitamin A (KU / kg)	0619	0.2685	0.0050	0.7876	0.6831	0.2725	6	-0.76	106.02	0
106	Vitamin A (KU / kg)	0676	0.3000	0.0000	0.7876	0.6831	0.2725	6	-0.71	106.02	0
106	Vitamin A (KU / kg)	0148	0.4850	0.2300	0.7876	0.6831	0.2725	6	-0.44	106.02	0
106	Vitamin A (KU / kg)	0171	0.6800	0.0400	0.7876	0.6831	0.2725	6	-0.16	106.00	0
106	Vitamin A (KU / kg)	0905	1.180	0.0000	0.7876	0.6831	0.2725	6	0.57	106.02	0
106	Vitamin A (KU / kg)	0038	4.440	1.360	0.7876	0.6831	0.2725	6	5.35	106.02	0
106	Vitamin A (KU / kg)	0675	7.935	2.210	0.7876	0.6831	0.2725	6	10.46	106.02	1

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
106	Vitamin A (KU / kg)	0199	< 0.2		0.7876	0.6831	0.2725	6		106.02	5
106	Vitamin A (KU / kg)	0098	< 0.8		0.7876	0.6831	0.2725	6		106.01	5
106	Vitamin A (KU / kg)	0910	< 1		0.7876	0.6831	0.2725	6		106.02	5
106	Vitamin A (KU / kg)	2053	< 1		0.7876	0.6831	0.2725	6		106.02	5
107	Vitamin B12 (µg / kg (ppb))	0227	< 0.004					0		107.00	5
108	Vitamin D3 (KU / kg)	0619	0.0745	0.0030				3		108.02	0
108	Vitamin D3 (KU / kg)	0675	1.440	0.0400				3		108.02	0
108	Vitamin D3 (KU / kg)	0676	3.340	0.2400				3		108.02	0
108	Vitamin D3 (KU / kg)	0910	< 0.5					3		108.02	5
109	Vitamin E (IU / kg)	0610	19.00	0.0000	24.77	6.297	2.011	12	-0.92	109.02	0
109	Vitamin E (IU / kg)	0910	19.45	2.500	24.77	6.297	2.011	12	-0.84	109.02	0
109	Vitamin E (IU / kg)	0676	20.43	0.5100	24.77	6.297	2.011	12	-0.69	109.02	0
109	Vitamin E (IU / kg)	0227	22.05	1.500	24.77	6.297	2.011	12	-0.43	109.02	0
109	Vitamin E (IU / kg)	0563	22.13	1.040	24.77	6.297	2.011	12	-0.42	109.02	0
109	Vitamin E (IU / kg)	0036	22.15	0.9000	24.77	6.297	2.011	12	-0.42	109.02	0
109	Vitamin E (IU / kg)	0199	22.70	1.030	24.77	6.297	2.011	12	-0.33	109.02	0
109	Vitamin E (IU / kg)	0619	23.00	2.000	24.77	6.297	2.011	12	-0.28	109.02	0
109	Vitamin E (IU / kg)	0148	26.07	5.540	24.77	6.297	2.011	12	0.21	109.02	0
109	Vitamin E (IU / kg)	0905	31.84	2.570	24.77	6.297	2.011	12	1.12	109.02	0
109	Vitamin E (IU / kg)	0675	54.55	6.240	24.77	6.297	2.011	12	4.73	109.02	0
109	Vitamin E (IU / kg)	0590	56.45	0.3000	24.77	6.297	2.011	12	5.03	109.02	0
109	Vitamin E (IU / kg)	0904	< 5		24.77	6.297	2.011	12		109.99	5
109	Vitamin E (IU / kg)	0098	< 40		24.77	6.297	2.011	12		109.02	5
112	Pyridoxine (µg / g)	0010	2.100	0.8000				2		112.01	0
112	Pyridoxine (µg / g)	0939	4.995	0.2100				2		112.01	0
114	Biotin (mg / kg (ppm))	0227	296.8	592.4				1		114.01	0
118	Peroxide value (meq/kg)	2081	4.000	2.640				1		118.99	0
120	Alanine (%)	0353	1.285	0.0700	1.494	0.0716	0.0240	30	-2.92	120.00	0
120	Alanine (%)	0675	1.290	0.0200	1.494	0.0716	0.0240	30	-2.85	120.00	0
120	Alanine (%)	2246	1.395	0.0700	1.494	0.0716	0.0240	30	-1.39	120.05	0
120	Alanine (%)	0889	1.414	0.0138	1.494	0.0716	0.0240	30	-1.12	120.99	0
120	Alanine (%)	0968	1.425	0.0010	1.494	0.0716	0.0240	30	-0.97	120.00	0
120	Alanine (%)	0610	1.425	0.0100	1.494	0.0716	0.0240	30	-0.97	120.05	0
120	Alanine (%)	0918	1.428	0.0407	1.494	0.0716	0.0240	30	-0.92	120.00	0
120	Alanine (%)	0910	1.430	0.0400	1.494	0.0716	0.0240	30	-0.90	120.00	0
120	Alanine (%)	0626	1.444	0.0010	1.494	0.0716	0.0240	30	-0.71	120.05	0
120	Alanine (%)	0652	1.470	0.0400	1.494	0.0716	0.0240	30	-0.34	120.00	0
120	Alanine (%)	0684	1.470	0.0200	1.494	0.0716	0.0240	30	-0.34	120.00	0
120	Alanine (%)	0676	1.475	0.0100	1.494	0.0716	0.0240	30	-0.27	120.05	0
120	Alanine (%)	2059	1.481	0.0120	1.494	0.0716	0.0240	30	-0.18	120.00	0
120	Alanine (%)	0682	1.501	0.0000	1.494	0.0716	0.0240	30	0.10	120.00	0
120	Alanine (%)	0571	1.502	0.0170	1.494	0.0716	0.0240	30	0.10	120.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
120	Alanine (%)	0644	1.503	0.0080	1.494	0.0716	0.0240	30	0.12	120.00	0
120	Alanine (%)	0872	1.512	0.0130	1.494	0.0716	0.0240	30	0.24	120.00	0
120	Alanine (%)	0226	1.520	0.1396	1.494	0.0716	0.0240	30	0.36	120.00	0
120	Alanine (%)	0354	1.523	0.0080	1.494	0.0716	0.0240	30	0.40	120.00	0
120	Alanine (%)	0868	1.524	0.0010	1.494	0.0716	0.0240	30	0.41	120.00	0
120	Alanine (%)	0504	1.525	0.0100	1.494	0.0716	0.0240	30	0.43	120.00	0
120	Alanine (%)	0939	1.535	0.0100	1.494	0.0716	0.0240	30	0.57	120.00	0
120	Alanine (%)	0227	1.540	0.0200	1.494	0.0716	0.0240	30	0.64	120.00	0
120	Alanine (%)	0870	1.546	0.0095	1.494	0.0716	0.0240	30	0.73	120.00	0
120	Alanine (%)	0098	1.557	0.0050	1.494	0.0716	0.0240	30	0.87	120.02	0
120	Alanine (%)	0859	1.558	0.0097	1.494	0.0716	0.0240	30	0.89	120.00	0
120	Alanine (%)	2188	1.570	0.0000	1.494	0.0716	0.0240	30	1.06	120.05	0
120	Alanine (%)	2196	1.585	0.0000	1.494	0.0716	0.0240	30	1.27	120.05	0
120	Alanine (%)	0619	1.595	0.0700	1.494	0.0716	0.0240	30	1.41	120.00	0
120	Alanine (%)	0148	1.635	0.0500	1.494	0.0716	0.0240	30	1.97	120.05	0
120	Alanine (%)	0407	1.728	0.1910	1.494	0.0716	0.0240	30	3.26	120.05	1
121	Arginine (%)	0889	1.492	0.0320	2.595	0.1146	0.0384	29	-9.63	121.99	0
121	Arginine (%)	0675	2.220	0.0200	2.595	0.1146	0.0384	29	-3.27	121.00	0
121	Arginine (%)	0868	2.377	0.0190	2.595	0.1146	0.0384	29	-1.91	121.00	0
121	Arginine (%)	0918	2.383	0.0218	2.595	0.1146	0.0384	29	-1.85	121.00	0
121	Arginine (%)	0968	2.484	0.0510	2.595	0.1146	0.0384	29	-0.97	121.00	0
121	Arginine (%)	0098	2.519	0.0080	2.595	0.1146	0.0384	29	-0.66	121.02	0
121	Arginine (%)	0353	2.525	0.0300	2.595	0.1146	0.0384	29	-0.61	121.00	0
121	Arginine (%)	0939	2.540	0.0600	2.595	0.1146	0.0384	29	-0.48	121.00	0
121	Arginine (%)	0226	2.544	0.2441	2.595	0.1146	0.0384	29	-0.45	121.00	0
121	Arginine (%)	0227	2.560	0.0400	2.595	0.1146	0.0384	29	-0.31	121.00	0
121	Arginine (%)	0682	2.564	0.0000	2.595	0.1146	0.0384	29	-0.27	121.00	0
121	Arginine (%)	0354	2.591	0.0190	2.595	0.1146	0.0384	29	-0.04	121.00	0
121	Arginine (%)	0626	2.591	0.0450	2.595	0.1146	0.0384	29	-0.04	121.05	0
121	Arginine (%)	0870	2.599	0.0146	2.595	0.1146	0.0384	29	0.03	121.00	0
121	Arginine (%)	0684	2.600	0.0600	2.595	0.1146	0.0384	29	0.04	121.00	0
121	Arginine (%)	0644	2.613	0.0110	2.595	0.1146	0.0384	29	0.15	121.00	0
121	Arginine (%)	2059	2.620	0.0010	2.595	0.1146	0.0384	29	0.21	121.00	0
121	Arginine (%)	0610	2.620	0.0000	2.595	0.1146	0.0384	29	0.22	121.05	0
121	Arginine (%)	0504	2.625	0.0300	2.595	0.1146	0.0384	29	0.26	121.00	0
121	Arginine (%)	0872	2.639	0.0210	2.595	0.1146	0.0384	29	0.38	121.00	0
121	Arginine (%)	0652	2.640	0.0400	2.595	0.1146	0.0384	29	0.39	121.00	0
121	Arginine (%)	0910	2.650	0.1800	2.595	0.1146	0.0384	29	0.48	121.00	0
121	Arginine (%)	0859	2.655	0.0302	2.595	0.1146	0.0384	29	0.52	121.00	0
121	Arginine (%)	0571	2.664	0.0160	2.595	0.1146	0.0384	29	0.60	121.00	0
121	Arginine (%)	2196	2.675	0.0000	2.595	0.1146	0.0384	29	0.70	121.05	0
121	Arginine (%)	2188	2.750	0.0000	2.595	0.1146	0.0384	29	1.35	121.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
121	Arginine (%)	0676	2.795	0.0100	2.595	0.1146	0.0384	29	1.74	121.05	0
121	Arginine (%)	0619	2.810	0.1000	2.595	0.1146	0.0384	29	1.88	121.00	0
121	Arginine (%)	0148	3.025	0.0100	2.595	0.1146	0.0384	29	3.75	121.05	0
121	Arginine (%)	0407	2.732	0.3510	2.595	0.1146	0.0384	29	1.19	121.05	1
122	Aspartic (%)	0889	2.154	0.0183	3.136	0.1206	0.0474	30	-8.14	122.99	0
122	Aspartic (%)	0407	2.496	0.1540	3.136	0.1206	0.0474	30	-5.30	122.05	0
122	Aspartic (%)	0675	2.510	0.0200	3.136	0.1206	0.0474	30	-5.19	122.00	0
122	Aspartic (%)	0676	2.910	0.1000	3.136	0.1206	0.0474	30	-1.87	122.05	0
122	Aspartic (%)	0353	2.950	0.0600	3.136	0.1206	0.0474	30	-1.54	122.00	0
122	Aspartic (%)	0610	3.065	0.0100	3.136	0.1206	0.0474	30	-0.59	122.05	0
122	Aspartic (%)	0682	3.071	0.0000	3.136	0.1206	0.0474	30	-0.54	122.00	0
122	Aspartic (%)	2059	3.087	0.0200	3.136	0.1206	0.0474	30	-0.40	122.00	0
122	Aspartic (%)	2246	3.095	0.1100	3.136	0.1206	0.0474	30	-0.34	122.05	0
122	Aspartic (%)	0354	3.103	0.0120	3.136	0.1206	0.0474	30	-0.27	122.00	0
122	Aspartic (%)	0684	3.120	0.0600	3.136	0.1206	0.0474	30	-0.13	122.00	0
122	Aspartic (%)	0644	3.134	0.0140	3.136	0.1206	0.0474	30	-0.01	122.00	0
122	Aspartic (%)	0872	3.135	0.0070	3.136	0.1206	0.0474	30	-0.01	122.00	0
122	Aspartic (%)	0968	3.135	0.0100	3.136	0.1206	0.0474	30	0.00	122.00	0
122	Aspartic (%)	0910	3.140	0.2200	3.136	0.1206	0.0474	30	0.04	122.00	0
122	Aspartic (%)	0868	3.149	0.0200	3.136	0.1206	0.0474	30	0.11	122.00	0
122	Aspartic (%)	0227	3.150	0.0400	3.136	0.1206	0.0474	30	0.12	122.00	0
122	Aspartic (%)	0098	3.155	0.0230	3.136	0.1206	0.0474	30	0.16	122.02	0
122	Aspartic (%)	0571	3.156	0.0390	3.136	0.1206	0.0474	30	0.17	122.00	0
122	Aspartic (%)	0939	3.160	0.0200	3.136	0.1206	0.0474	30	0.20	122.00	0
122	Aspartic (%)	0504	3.175	0.0500	3.136	0.1206	0.0474	30	0.33	122.00	0
122	Aspartic (%)	0626	3.183	0.0060	3.136	0.1206	0.0474	30	0.39	122.05	0
122	Aspartic (%)	0652	3.195	0.0100	3.136	0.1206	0.0474	30	0.49	122.00	0
122	Aspartic (%)	0859	3.197	0.0321	3.136	0.1206	0.0474	30	0.51	122.00	0
122	Aspartic (%)	0918	3.220	0.0870	3.136	0.1206	0.0474	30	0.70	122.00	0
122	Aspartic (%)	2188	3.240	0.0000	3.136	0.1206	0.0474	30	0.87	122.05	0
122	Aspartic (%)	0870	3.279	0.0393	3.136	0.1206	0.0474	30	1.19	122.00	0
122	Aspartic (%)	2196	3.388	0.0000	3.136	0.1206	0.0474	30	2.09	122.05	0
122	Aspartic (%)	0619	3.400	0.0400	3.136	0.1206	0.0474	30	2.19	122.00	0
122	Aspartic (%)	0148	3.520	0.2000	3.136	0.1206	0.0474	30	3.19	122.05	0
122	Aspartic (%)	0226	3.189	0.3052	3.136	0.1206	0.0474	30	0.44	122.00	1
124	Cysteine/Cystine (%)	0610	0.2200	0.0000	0.5001	0.0434	0.0146	29	-6.45	124.05	0
124	Cysteine/Cystine (%)	2246	0.3300	0.0200	0.5001	0.0434	0.0146	29	-3.92	124.05	0
124	Cysteine/Cystine (%)	2196	0.4190	0.0000	0.5001	0.0434	0.0146	29	-1.87	124.05	0
124	Cysteine/Cystine (%)	0870	0.4578	0.0080	0.5001	0.0434	0.0146	29	-0.97	124.00	0
124	Cysteine/Cystine (%)	0353	0.4600	0.0000	0.5001	0.0434	0.0146	29	-0.92	124.00	0
124	Cysteine/Cystine (%)	0918	0.4650	0.0500	0.5001	0.0434	0.0146	29	-0.81	124.00	0
124	Cysteine/Cystine (%)	0227	0.4700	0.0400	0.5001	0.0434	0.0146	29	-0.69	124.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
124	Cysteine/Cystine (%)	0910	0.4750	0.0100	0.5001	0.0434	0.0146	29	-0.58	124.00	0
124	Cysteine/Cystine (%)	0644	0.4775	0.0050	0.5001	0.0434	0.0146	29	-0.52	124.00	0
124	Cysteine/Cystine (%)	0872	0.4810	0.0120	0.5001	0.0434	0.0146	29	-0.44	124.00	0
124	Cysteine/Cystine (%)	0652	0.4850	0.0100	0.5001	0.0434	0.0146	29	-0.35	124.00	0
124	Cysteine/Cystine (%)	2059	0.4865	0.0010	0.5001	0.0434	0.0146	29	-0.31	124.00	0
124	Cysteine/Cystine (%)	0354	0.4895	0.0250	0.5001	0.0434	0.0146	29	-0.24	124.00	0
124	Cysteine/Cystine (%)	0868	0.4975	0.0030	0.5001	0.0434	0.0146	29	-0.06	124.00	0
124	Cysteine/Cystine (%)	0684	0.5000	0.0200	0.5001	0.0434	0.0146	29	0.00	124.00	0
124	Cysteine/Cystine (%)	0226	0.5066	0.0204	0.5001	0.0434	0.0146	29	0.15	124.00	0
124	Cysteine/Cystine (%)	0098	0.5075	0.0230	0.5001	0.0434	0.0146	29	0.17	124.02	0
124	Cysteine/Cystine (%)	0968	0.5090	0.0060	0.5001	0.0434	0.0146	29	0.21	124.00	0
124	Cysteine/Cystine (%)	0571	0.5110	0.0140	0.5001	0.0434	0.0146	29	0.25	124.00	0
124	Cysteine/Cystine (%)	2188	0.5150	0.0100	0.5001	0.0434	0.0146	29	0.34	124.05	0
124	Cysteine/Cystine (%)	0889	0.5247	0.0046	0.5001	0.0434	0.0146	29	0.57	124.99	0
124	Cysteine/Cystine (%)	0939	0.5250	0.0100	0.5001	0.0434	0.0146	29	0.57	124.00	0
124	Cysteine/Cystine (%)	0859	0.5314	0.0121	0.5001	0.0434	0.0146	29	0.72	124.00	0
124	Cysteine/Cystine (%)	0682	0.5370	0.0000	0.5001	0.0434	0.0146	29	0.85	124.00	0
124	Cysteine/Cystine (%)	0675	0.5400	0.0200	0.5001	0.0434	0.0146	29	0.92	124.00	0
124	Cysteine/Cystine (%)	0504	0.5550	0.0300	0.5001	0.0434	0.0146	29	1.26	124.00	0
124	Cysteine/Cystine (%)	0676	0.5600	0.0000	0.5001	0.0434	0.0146	29	1.38	124.05	0
124	Cysteine/Cystine (%)	0148	0.6000	0.0400	0.5001	0.0434	0.0146	29	2.30	124.05	0
124	Cysteine/Cystine (%)	0619	0.6255	0.0290	0.5001	0.0434	0.0146	29	2.89	124.00	0
124	Cysteine/Cystine (%)	0407	0.5095	0.2870	0.5001	0.0434	0.0146	29	0.22	124.05	1
125	Glutamic (%)	0889	4.674	0.0733	5.922	0.2759	0.1052	30	-4.52	125.99	0
125	Glutamic (%)	0675	4.810	0.0200	5.922	0.2759	0.1052	30	-4.03	125.00	0
125	Glutamic (%)	2196	5.399	0.0000	5.922	0.2759	0.1052	30	-1.89	125.05	0
125	Glutamic (%)	0868	5.659	0.0630	5.922	0.2759	0.1052	30	-0.95	125.00	0
125	Glutamic (%)	0652	5.705	0.2300	5.922	0.2759	0.1052	30	-0.78	125.00	0
125	Glutamic (%)	0098	5.707	0.1710	5.922	0.2759	0.1052	30	-0.78	125.02	0
125	Glutamic (%)	2059	5.709	0.0110	5.922	0.2759	0.1052	30	-0.77	125.00	0
125	Glutamic (%)	0610	5.740	0.1200	5.922	0.2759	0.1052	30	-0.66	125.05	0
125	Glutamic (%)	0872	5.807	0.0220	5.922	0.2759	0.1052	30	-0.42	125.00	0
125	Glutamic (%)	0353	5.815	0.0300	5.922	0.2759	0.1052	30	-0.39	125.00	0
125	Glutamic (%)	0504	5.815	0.1100	5.922	0.2759	0.1052	30	-0.39	125.00	0
125	Glutamic (%)	0684	5.815	0.0500	5.922	0.2759	0.1052	30	-0.39	125.00	0
125	Glutamic (%)	0571	5.875	0.0910	5.922	0.2759	0.1052	30	-0.17	125.00	0
125	Glutamic (%)	2188	5.890	0.0000	5.922	0.2759	0.1052	30	-0.11	125.05	0
125	Glutamic (%)	0148	5.905	0.3500	5.922	0.2759	0.1052	30	-0.06	125.05	0
125	Glutamic (%)	0968	5.906	0.0710	5.922	0.2759	0.1052	30	-0.06	125.00	0
125	Glutamic (%)	0682	5.934	0.0000	5.922	0.2759	0.1052	30	0.05	125.00	0
125	Glutamic (%)	0227	5.935	0.0500	5.922	0.2759	0.1052	30	0.05	125.00	0
125	Glutamic (%)	0644	5.953	0.0100	5.922	0.2759	0.1052	30	0.11	125.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
125	Glutamic (%)	0859	5.964	0.0788	5.922	0.2759	0.1052	30	0.16	125.00	0
125	Glutamic (%)	0676	5.985	0.1500	5.922	0.2759	0.1052	30	0.23	125.05	0
125	Glutamic (%)	0910	6.015	0.3900	5.922	0.2759	0.1052	30	0.34	125.00	0
125	Glutamic (%)	0354	6.143	0.0100	5.922	0.2759	0.1052	30	0.80	125.00	0
125	Glutamic (%)	0626	6.145	0.0110	5.922	0.2759	0.1052	30	0.81	125.05	0
125	Glutamic (%)	0918	6.243	0.0022	5.922	0.2759	0.1052	30	1.16	125.00	0
125	Glutamic (%)	0870	6.244	0.0547	5.922	0.2759	0.1052	30	1.17	125.00	0
125	Glutamic (%)	0226	6.271	0.6069	5.922	0.2759	0.1052	30	1.27	125.00	0
125	Glutamic (%)	2246	6.275	0.1700	5.922	0.2759	0.1052	30	1.28	125.05	0
125	Glutamic (%)	0619	6.425	0.1700	5.922	0.2759	0.1052	30	1.82	125.00	0
125	Glutamic (%)	0939	6.490	0.0400	5.922	0.2759	0.1052	30	2.06	125.00	0
125	Glutamic (%)	0407	6.469	0.9450	5.922	0.2759	0.1052	30	1.98	125.05	1
126	Glycine (%)	0889	0.5201	0.0046	1.613	0.0874	0.0240	30	-12.50	126.99	0
126	Glycine (%)	0675	1.380	0.0200	1.613	0.0874	0.0240	30	-2.67	126.00	0
126	Glycine (%)	2246	1.435	0.0900	1.613	0.0874	0.0240	30	-2.04	126.05	0
126	Glycine (%)	2196	1.471	0.0000	1.613	0.0874	0.0240	30	-1.63	126.05	0
126	Glycine (%)	0353	1.475	0.0500	1.613	0.0874	0.0240	30	-1.58	126.00	0
126	Glycine (%)	0652	1.575	0.0500	1.613	0.0874	0.0240	30	-0.44	126.00	0
126	Glycine (%)	0676	1.575	0.0100	1.613	0.0874	0.0240	30	-0.44	126.05	0
126	Glycine (%)	0968	1.577	0.0000	1.613	0.0874	0.0240	30	-0.42	126.00	0
126	Glycine (%)	0684	1.585	0.0300	1.613	0.0874	0.0240	30	-0.32	126.00	0
126	Glycine (%)	0910	1.585	0.0900	1.613	0.0874	0.0240	30	-0.32	126.00	0
126	Glycine (%)	2059	1.599	0.0070	1.613	0.0874	0.0240	30	-0.17	126.00	0
126	Glycine (%)	0682	1.600	0.0000	1.613	0.0874	0.0240	30	-0.15	126.00	0
126	Glycine (%)	0610	1.605	0.0100	1.613	0.0874	0.0240	30	-0.10	126.05	0
126	Glycine (%)	0354	1.611	0.0010	1.613	0.0874	0.0240	30	-0.03	126.00	0
126	Glycine (%)	0872	1.622	0.0100	1.613	0.0874	0.0240	30	0.10	126.00	0
126	Glycine (%)	0098	1.623	0.0180	1.613	0.0874	0.0240	30	0.11	126.02	0
126	Glycine (%)	0918	1.624	0.0233	1.613	0.0874	0.0240	30	0.12	126.00	0
126	Glycine (%)	0868	1.626	0.0100	1.613	0.0874	0.0240	30	0.14	126.00	0
126	Glycine (%)	0571	1.629	0.0150	1.613	0.0874	0.0240	30	0.17	126.00	0
126	Glycine (%)	0644	1.633	0.0090	1.613	0.0874	0.0240	30	0.22	126.00	0
126	Glycine (%)	0939	1.640	0.0000	1.613	0.0874	0.0240	30	0.30	126.00	0
126	Glycine (%)	0227	1.650	0.0200	1.613	0.0874	0.0240	30	0.42	126.00	0
126	Glycine (%)	0859	1.656	0.0054	1.613	0.0874	0.0240	30	0.48	126.00	0
126	Glycine (%)	0504	1.660	0.0200	1.613	0.0874	0.0240	30	0.53	126.00	0
126	Glycine (%)	0870	1.668	0.0131	1.613	0.0874	0.0240	30	0.62	126.00	0
126	Glycine (%)	2188	1.690	0.0000	1.613	0.0874	0.0240	30	0.88	126.05	0
126	Glycine (%)	0626	1.731	0.0150	1.613	0.0874	0.0240	30	1.34	126.05	0
126	Glycine (%)	0619	1.740	0.0600	1.613	0.0874	0.0240	30	1.45	126.00	0
126	Glycine (%)	0148	1.855	0.1300	1.613	0.0874	0.0240	30	2.76	126.05	0
126	Glycine (%)	0407	2.115	0.0080	1.613	0.0874	0.0240	30	5.74	126.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
126	Glycine (%)	0226	1.597	0.1874	1.613	0.0874	0.0240	30	-0.19	126.00	1
127	Histidine (%)	0889	0.5201	0.0046	0.8035	0.0564	0.0217	31	-5.02	127.99	0
127	Histidine (%)	0407	0.5410	0.0880	0.8035	0.0564	0.0217	31	-4.65	127.05	0
127	Histidine (%)	0675	0.6400	0.0200	0.8035	0.0564	0.0217	31	-2.90	127.00	0
127	Histidine (%)	0353	0.6900	0.0200	0.8035	0.0564	0.0217	31	-2.01	127.00	0
127	Histidine (%)	2246	0.7150	0.0500	0.8035	0.0564	0.0217	31	-1.57	127.05	0
127	Histidine (%)	0226	0.7459	0.0640	0.8035	0.0564	0.0217	31	-1.02	127.00	0
127	Histidine (%)	0626	0.7690	0.0120	0.8035	0.0564	0.0217	31	-0.61	127.05	0
127	Histidine (%)	0676	0.7750	0.0500	0.8035	0.0564	0.0217	31	-0.51	127.05	0
127	Histidine (%)	2188	0.7800	0.0000	0.8035	0.0564	0.0217	31	-0.42	127.05	0
127	Histidine (%)	0354	0.7810	0.0100	0.8035	0.0564	0.0217	31	-0.40	127.00	0
127	Histidine (%)	0939	0.7950	0.0300	0.8035	0.0564	0.0217	31	-0.15	127.00	0
127	Histidine (%)	2196	0.8040	0.0000	0.8035	0.0564	0.0217	31	0.01	127.05	0
127	Histidine (%)	0968	0.8055	0.0010	0.8035	0.0564	0.0217	31	0.03	127.00	0
127	Histidine (%)	0098	0.8080	0.0000	0.8035	0.0564	0.0217	31	0.08	127.02	0
127	Histidine (%)	0684	0.8100	0.0400	0.8035	0.0564	0.0217	31	0.11	127.00	0
127	Histidine (%)	0910	0.8100	0.0400	0.8035	0.0564	0.0217	31	0.11	127.00	0
127	Histidine (%)	0610	0.8100	0.0200	0.8035	0.0564	0.0217	31	0.11	127.05	0
127	Histidine (%)	2059	0.8105	0.0010	0.8035	0.0564	0.0217	31	0.12	127.00	0
127	Histidine (%)	0872	0.8165	0.0050	0.8035	0.0564	0.0217	31	0.23	127.00	0
127	Histidine (%)	0571	0.8180	0.0080	0.8035	0.0564	0.0217	31	0.26	127.00	0
127	Histidine (%)	0870	0.8251	0.0169	0.8035	0.0564	0.0217	31	0.38	127.00	0
127	Histidine (%)	0644	0.8295	0.0010	0.8035	0.0564	0.0217	31	0.46	127.00	0
127	Histidine (%)	0859	0.8362	0.0048	0.8035	0.0564	0.0217	31	0.58	127.00	0
127	Histidine (%)	0682	0.8410	0.0000	0.8035	0.0564	0.0217	31	0.66	127.00	0
127	Histidine (%)	0868	0.8435	0.0310	0.8035	0.0564	0.0217	31	0.71	127.00	0
127	Histidine (%)	0227	0.8450	0.0100	0.8035	0.0564	0.0217	31	0.74	127.00	0
127	Histidine (%)	0918	0.8500	0.0119	0.8035	0.0564	0.0217	31	0.82	127.00	0
127	Histidine (%)	0652	0.8500	0.0400	0.8035	0.0564	0.0217	31	0.82	127.00	0
127	Histidine (%)	0504	0.8800	0.0000	0.8035	0.0564	0.0217	31	1.36	127.00	0
127	Histidine (%)	0148	0.9750	0.0700	0.8035	0.0564	0.0217	31	3.04	127.05	0
127	Histidine (%)	0619	0.9820	0.0220	0.8035	0.0564	0.0217	31	3.16	127.00	0
128	Isoleucine (%)	0353	0.8600	0.0400	1.176	0.0796	0.0281	30	-3.97	128.00	0
128	Isoleucine (%)	0675	0.9000	0.0200	1.176	0.0796	0.0281	30	-3.47	128.00	0
128	Isoleucine (%)	0889	0.9807	0.0091	1.176	0.0796	0.0281	30	-2.46	128.99	0
128	Isoleucine (%)	0918	1.079	0.0106	1.176	0.0796	0.0281	30	-1.22	128.00	0
128	Isoleucine (%)	0354	1.096	0.0190	1.176	0.0796	0.0281	30	-1.01	128.00	0
128	Isoleucine (%)	0939	1.100	0.0400	1.176	0.0796	0.0281	30	-0.96	128.00	0
128	Isoleucine (%)	2246	1.100	0.0200	1.176	0.0796	0.0281	30	-0.96	128.05	0
128	Isoleucine (%)	0684	1.110	0.1200	1.176	0.0796	0.0281	30	-0.83	128.00	0
128	Isoleucine (%)	0968	1.145	0.0250	1.176	0.0796	0.0281	30	-0.40	128.00	0
128	Isoleucine (%)	0682	1.148	0.0000	1.176	0.0796	0.0281	30	-0.36	128.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
128	Isoleucine (%)	0619	1.155	0.0100	1.176	0.0796	0.0281	30	-0.27	128.00	0
128	Isoleucine (%)	0676	1.160	0.0400	1.176	0.0796	0.0281	30	-0.20	128.05	0
128	Isoleucine (%)	0868	1.163	0.0720	1.176	0.0796	0.0281	30	-0.17	128.00	0
128	Isoleucine (%)	0870	1.172	0.0359	1.176	0.0796	0.0281	30	-0.05	128.00	0
128	Isoleucine (%)	0098	1.190	0.0030	1.176	0.0796	0.0281	30	0.17	128.02	0
128	Isoleucine (%)	0652	1.190	0.0000	1.176	0.0796	0.0281	30	0.17	128.00	0
128	Isoleucine (%)	0910	1.195	0.0700	1.176	0.0796	0.0281	30	0.24	128.00	0
128	Isoleucine (%)	0626	1.202	0.0570	1.176	0.0796	0.0281	30	0.32	128.05	0
128	Isoleucine (%)	0226	1.210	0.0242	1.176	0.0796	0.0281	30	0.42	128.00	0
128	Isoleucine (%)	2059	1.213	0.0130	1.176	0.0796	0.0281	30	0.45	128.00	0
128	Isoleucine (%)	0872	1.214	0.0430	1.176	0.0796	0.0281	30	0.47	128.00	0
128	Isoleucine (%)	0571	1.218	0.0340	1.176	0.0796	0.0281	30	0.52	128.00	0
128	Isoleucine (%)	0227	1.225	0.0300	1.176	0.0796	0.0281	30	0.61	128.00	0
128	Isoleucine (%)	0610	1.225	0.0100	1.176	0.0796	0.0281	30	0.61	128.05	0
128	Isoleucine (%)	2188	1.240	0.0000	1.176	0.0796	0.0281	30	0.80	128.05	0
128	Isoleucine (%)	2196	1.249	0.0000	1.176	0.0796	0.0281	30	0.91	128.05	0
128	Isoleucine (%)	0644	1.251	0.0000	1.176	0.0796	0.0281	30	0.94	128.00	0
128	Isoleucine (%)	0504	1.285	0.0500	1.176	0.0796	0.0281	30	1.37	128.00	0
128	Isoleucine (%)	0859	1.289	0.0158	1.176	0.0796	0.0281	30	1.42	128.00	0
128	Isoleucine (%)	0148	1.395	0.0300	1.176	0.0796	0.0281	30	2.75	128.05	0
128	Isoleucine (%)	0407	1.395	0.2850	1.176	0.0796	0.0281	30	2.74	128.05	1
129	Leucine (%)	0889	1.773	0.0000	2.328	0.0621	0.0345	30	-8.93	129.99	0
129	Leucine (%)	0353	1.985	0.0500	2.328	0.0621	0.0345	30	-5.52	129.00	0
129	Leucine (%)	0675	2.000	0.0200	2.328	0.0621	0.0345	30	-5.28	129.00	0
129	Leucine (%)	2196	2.201	0.0000	2.328	0.0621	0.0345	30	-2.04	129.05	0
129	Leucine (%)	0684	2.275	0.0700	2.328	0.0621	0.0345	30	-0.85	129.00	0
129	Leucine (%)	0910	2.290	0.1200	2.328	0.0621	0.0345	30	-0.60	129.00	0
129	Leucine (%)	0682	2.291	0.0000	2.328	0.0621	0.0345	30	-0.59	129.00	0
129	Leucine (%)	2059	2.300	0.0130	2.328	0.0621	0.0345	30	-0.45	129.00	0
129	Leucine (%)	2246	2.300	0.0400	2.328	0.0621	0.0345	30	-0.44	129.05	0
129	Leucine (%)	0652	2.305	0.0100	2.328	0.0621	0.0345	30	-0.36	129.00	0
129	Leucine (%)	0939	2.305	0.0500	2.328	0.0621	0.0345	30	-0.36	129.00	0
129	Leucine (%)	0870	2.305	0.0287	2.328	0.0621	0.0345	30	-0.36	129.00	0
129	Leucine (%)	2188	2.310	0.0000	2.328	0.0621	0.0345	30	-0.28	129.05	0
129	Leucine (%)	0918	2.315	0.0289	2.328	0.0621	0.0345	30	-0.20	129.00	0
129	Leucine (%)	0968	2.323	0.0540	2.328	0.0621	0.0345	30	-0.07	129.00	0
129	Leucine (%)	0098	2.325	0.0050	2.328	0.0621	0.0345	30	-0.05	129.02	0
129	Leucine (%)	0610	2.325	0.0300	2.328	0.0621	0.0345	30	-0.04	129.05	0
129	Leucine (%)	0571	2.331	0.0400	2.328	0.0621	0.0345	30	0.06	129.00	0
129	Leucine (%)	0868	2.354	0.0090	2.328	0.0621	0.0345	30	0.42	129.00	0
129	Leucine (%)	0872	2.356	0.0040	2.328	0.0621	0.0345	30	0.46	129.00	0
129	Leucine (%)	0676	2.360	0.0400	2.328	0.0621	0.0345	30	0.52	129.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
129	Leucine (%)	0644	2.367	0.0030	2.328	0.0621	0.0345	30	0.63	129.00	0
129	Leucine (%)	0504	2.370	0.0200	2.328	0.0621	0.0345	30	0.68	129.00	0
129	Leucine (%)	0354	2.371	0.0060	2.328	0.0621	0.0345	30	0.70	129.00	0
129	Leucine (%)	0626	2.377	0.0220	2.328	0.0621	0.0345	30	0.80	129.05	0
129	Leucine (%)	0227	2.380	0.0200	2.328	0.0621	0.0345	30	0.85	129.00	0
129	Leucine (%)	0226	2.392	0.1722	2.328	0.0621	0.0345	30	1.04	129.00	0
129	Leucine (%)	0859	2.434	0.0291	2.328	0.0621	0.0345	30	1.71	129.00	0
129	Leucine (%)	0619	2.445	0.0500	2.328	0.0621	0.0345	30	1.89	129.00	0
129	Leucine (%)	0148	2.640	0.1000	2.328	0.0621	0.0345	30	5.03	129.05	0
129	Leucine (%)	0407	2.696	0.3040	2.328	0.0621	0.0345	30	5.94	129.05	1
130	L-Lysine (%)	0675	1.170	0.0200	1.443	0.0761	0.0233	32	-3.59	130.00	0
130	L-Lysine (%)	0353	1.185	0.0100	1.443	0.0761	0.0233	32	-3.40	130.00	0
130	L-Lysine (%)	0848	1.260	0.0000	1.443	0.0761	0.0233	32	-2.41	130.00	0
130	L-Lysine (%)	2246	1.355	0.0900	1.443	0.0761	0.0233	32	-1.16	130.05	0
130	L-Lysine (%)	0652	1.390	0.0200	1.443	0.0761	0.0233	32	-0.70	130.00	0
130	L-Lysine (%)	0610	1.395	0.0100	1.443	0.0761	0.0233	32	-0.64	130.05	0
130	L-Lysine (%)	0354	1.397	0.0240	1.443	0.0761	0.0233	32	-0.61	130.00	0
130	L-Lysine (%)	0682	1.404	0.0000	1.443	0.0761	0.0233	32	-0.52	130.00	0
130	L-Lysine (%)	0684	1.405	0.0500	1.443	0.0761	0.0233	32	-0.50	130.00	0
130	L-Lysine (%)	0226	1.406	0.0997	1.443	0.0761	0.0233	32	-0.49	130.00	0
130	L-Lysine (%)	0872	1.409	0.0190	1.443	0.0761	0.0233	32	-0.46	130.00	0
130	L-Lysine (%)	0939	1.410	0.0200	1.443	0.0761	0.0233	32	-0.44	130.00	0
130	L-Lysine (%)	0910	1.410	0.0800	1.443	0.0761	0.0233	32	-0.44	130.00	0
130	L-Lysine (%)	0868	1.412	0.0310	1.443	0.0761	0.0233	32	-0.42	130.00	0
130	L-Lysine (%)	2059	1.419	0.0070	1.443	0.0761	0.0233	32	-0.33	130.00	0
130	L-Lysine (%)	0027	1.434	0.0070	1.443	0.0761	0.0233	32	-0.13	130.05	0
130	L-Lysine (%)	0644	1.438	0.0080	1.443	0.0761	0.0233	32	-0.07	130.00	0
130	L-Lysine (%)	0571	1.439	0.0240	1.443	0.0761	0.0233	32	-0.06	130.00	0
130	L-Lysine (%)	0407	1.439	0.0700	1.443	0.0761	0.0233	32	-0.06	130.05	0
130	L-Lysine (%)	2196	1.452	0.0000	1.443	0.0761	0.0233	32	0.11	130.05	0
130	L-Lysine (%)	2188	1.460	0.0000	1.443	0.0761	0.0233	32	0.22	130.05	0
130	L-Lysine (%)	0968	1.477	0.0100	1.443	0.0761	0.0233	32	0.44	130.00	0
130	L-Lysine (%)	0918	1.484	0.0185	1.443	0.0761	0.0233	32	0.53	130.00	0
130	L-Lysine (%)	0859	1.490	0.0159	1.443	0.0761	0.0233	32	0.61	130.00	0
130	L-Lysine (%)	0626	1.491	0.0040	1.443	0.0761	0.0233	32	0.63	130.05	0
130	L-Lysine (%)	0504	1.525	0.0300	1.443	0.0761	0.0233	32	1.07	130.00	0
130	L-Lysine (%)	0619	1.525	0.0300	1.443	0.0761	0.0233	32	1.07	130.00	0
130	L-Lysine (%)	0870	1.525	0.0139	1.443	0.0761	0.0233	32	1.08	130.00	0
130	L-Lysine (%)	0098	1.540	0.0050	1.443	0.0761	0.0233	32	1.26	130.02	0
130	L-Lysine (%)	0148	1.615	0.0100	1.443	0.0761	0.0233	32	2.25	130.05	0
130	L-Lysine (%)	0227	1.635	0.0100	1.443	0.0761	0.0233	32	2.52	130.00	0
130	L-Lysine (%)	0889	1.737	0.0092	1.443	0.0761	0.0233	32	3.85	130.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
130	L-Lysine (%)	0676	2.010	0.9800	1.443	0.0761	0.0233	32	7.44	130.05	1
131	Methionine (%)	0610	0.1800	0.0000	0.4494	0.0399	0.0163	32	-6.76	131.05	0
131	Methionine (%)	2246	0.2400	0.0200	0.4494	0.0399	0.0163	32	-5.25	131.05	0
131	Methionine (%)	0675	0.3500	0.0200	0.4494	0.0399	0.0163	32	-2.49	131.00	0
131	Methionine (%)	0626	0.3515	0.0090	0.4494	0.0399	0.0163	32	-2.46	131.05	0
131	Methionine (%)	0148	0.4050	0.0300	0.4494	0.0399	0.0163	32	-1.11	131.05	0
131	Methionine (%)	2188	0.4200	0.0000	0.4494	0.0399	0.0163	32	-0.74	131.05	0
131	Methionine (%)	0684	0.4250	0.0100	0.4494	0.0399	0.0163	32	-0.61	131.00	0
131	Methionine (%)	0939	0.4300	0.0200	0.4494	0.0399	0.0163	32	-0.49	131.00	0
131	Methionine (%)	0027	0.4325	0.0510	0.4494	0.0399	0.0163	32	-0.42	131.05	0
131	Methionine (%)	0682	0.4340	0.0000	0.4494	0.0399	0.0163	32	-0.39	131.00	0
131	Methionine (%)	0918	0.4350	0.0500	0.4494	0.0399	0.0163	32	-0.36	131.00	0
131	Methionine (%)	0098	0.4355	0.0110	0.4494	0.0399	0.0163	32	-0.35	131.02	0
131	Methionine (%)	0968	0.4360	0.0200	0.4494	0.0399	0.0163	32	-0.34	131.00	0
131	Methionine (%)	0354	0.4375	0.0270	0.4494	0.0399	0.0163	32	-0.30	131.00	0
131	Methionine (%)	0652	0.4500	0.0200	0.4494	0.0399	0.0163	32	0.01	131.00	0
131	Methionine (%)	2059	0.4500	0.0020	0.4494	0.0399	0.0163	32	0.01	131.00	0
131	Methionine (%)	0872	0.4525	0.0110	0.4494	0.0399	0.0163	32	0.08	131.00	0
131	Methionine (%)	0407	0.4545	0.0650	0.4494	0.0399	0.0163	32	0.13	131.05	0
131	Methionine (%)	0226	0.4548	0.0072	0.4494	0.0399	0.0163	32	0.13	131.00	0
131	Methionine (%)	0504	0.4550	0.0100	0.4494	0.0399	0.0163	32	0.14	131.00	0
131	Methionine (%)	0571	0.4615	0.0090	0.4494	0.0399	0.0163	32	0.30	131.00	0
131	Methionine (%)	0868	0.4645	0.0330	0.4494	0.0399	0.0163	32	0.38	131.00	0
131	Methionine (%)	0353	0.4650	0.0300	0.4494	0.0399	0.0163	32	0.39	131.00	0
131	Methionine (%)	0644	0.4680	0.0120	0.4494	0.0399	0.0163	32	0.47	131.00	0
131	Methionine (%)	0619	0.4710	0.0080	0.4494	0.0399	0.0163	32	0.54	131.00	0
131	Methionine (%)	0227	0.4750	0.0100	0.4494	0.0399	0.0163	32	0.64	131.00	0
131	Methionine (%)	2196	0.4840	0.0000	0.4494	0.0399	0.0163	32	0.87	131.05	0
131	Methionine (%)	0848	0.4900	0.0000	0.4494	0.0399	0.0163	32	1.02	131.00	0
131	Methionine (%)	0676	0.5200	0.0200	0.4494	0.0399	0.0163	32	1.77	131.05	0
131	Methionine (%)	0870	0.5225	0.0059	0.4494	0.0399	0.0163	32	1.83	131.00	0
131	Methionine (%)	0859	0.5606	0.0056	0.4494	0.0399	0.0163	32	2.79	131.00	0
131	Methionine (%)	0889	0.5614	0.0045	0.4494	0.0399	0.0163	32	2.81	131.99	0
131	Methionine (%)	0910	0.4950	0.0900	0.4494	0.0399	0.0163	32	1.14	131.00	1
132	Phenylalanine (%)	0889	1.180	0.0138	1.540	0.0902	0.0297	30	-3.99	132.99	0
132	Phenylalanine (%)	0675	1.240	0.0200	1.540	0.0902	0.0297	30	-3.32	132.00	0
132	Phenylalanine (%)	2196	1.380	0.0000	1.540	0.0902	0.0297	30	-1.77	132.05	0
132	Phenylalanine (%)	2246	1.410	0.1400	1.540	0.0902	0.0297	30	-1.44	132.05	0
132	Phenylalanine (%)	0682	1.433	0.0000	1.540	0.0902	0.0297	30	-1.19	132.00	0
132	Phenylalanine (%)	0353	1.465	0.0300	1.540	0.0902	0.0297	30	-0.83	132.00	0
132	Phenylalanine (%)	0684	1.485	0.0500	1.540	0.0902	0.0297	30	-0.61	132.00	0
132	Phenylalanine (%)	0610	1.505	0.0100	1.540	0.0902	0.0297	30	-0.39	132.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
132	Phenylalanine (%)	0354	1.509	0.0070	1.540	0.0902	0.0297	30	-0.35	132.00	0
132	Phenylalanine (%)	0652	1.510	0.0600	1.540	0.0902	0.0297	30	-0.33	132.00	0
132	Phenylalanine (%)	0939	1.510	0.0400	1.540	0.0902	0.0297	30	-0.33	132.00	0
132	Phenylalanine (%)	2059	1.513	0.0050	1.540	0.0902	0.0297	30	-0.30	132.00	0
132	Phenylalanine (%)	0098	1.516	0.0120	1.540	0.0902	0.0297	30	-0.27	132.02	0
132	Phenylalanine (%)	0910	1.550	0.1200	1.540	0.0902	0.0297	30	0.11	132.00	0
132	Phenylalanine (%)	0571	1.555	0.0200	1.540	0.0902	0.0297	30	0.17	132.00	0
132	Phenylalanine (%)	0870	1.556	0.0051	1.540	0.0902	0.0297	30	0.18	132.00	0
132	Phenylalanine (%)	0872	1.558	0.0230	1.540	0.0902	0.0297	30	0.19	132.00	0
132	Phenylalanine (%)	0676	1.565	0.0500	1.540	0.0902	0.0297	30	0.28	132.05	0
132	Phenylalanine (%)	0918	1.566	0.0238	1.540	0.0902	0.0297	30	0.29	132.00	0
132	Phenylalanine (%)	0227	1.570	0.0200	1.540	0.0902	0.0297	30	0.33	132.00	0
132	Phenylalanine (%)	0868	1.571	0.0500	1.540	0.0902	0.0297	30	0.34	132.00	0
132	Phenylalanine (%)	0644	1.572	0.0060	1.540	0.0902	0.0297	30	0.36	132.00	0
132	Phenylalanine (%)	0626	1.578	0.0300	1.540	0.0902	0.0297	30	0.42	132.05	0
132	Phenylalanine (%)	2188	1.590	0.0000	1.540	0.0902	0.0297	30	0.55	132.05	0
132	Phenylalanine (%)	0968	1.599	0.0210	1.540	0.0902	0.0297	30	0.65	132.00	0
132	Phenylalanine (%)	0619	1.630	0.0600	1.540	0.0902	0.0297	30	1.00	132.00	0
132	Phenylalanine (%)	0504	1.645	0.0100	1.540	0.0902	0.0297	30	1.16	132.00	0
132	Phenylalanine (%)	0859	1.745	0.0148	1.540	0.0902	0.0297	30	2.27	132.00	0
132	Phenylalanine (%)	0148	1.815	0.0100	1.540	0.0902	0.0297	30	3.05	132.05	0
132	Phenylalanine (%)	0407	1.844	0.0390	1.540	0.0902	0.0297	30	3.36	132.05	0
132	Phenylalanine (%)	0226	1.481	0.1571	1.540	0.0902	0.0297	30	-0.65	132.00	1
133	Proline (%)	0682	1.202	0.0000	1.756	0.1771	0.0393	29	-3.13	133.00	0
133	Proline (%)	0675	1.430	0.0200	1.756	0.1771	0.0393	29	-1.84	133.00	0
133	Proline (%)	2196	1.474	0.0000	1.756	0.1771	0.0393	29	-1.59	133.05	0
133	Proline (%)	0889	1.574	0.0046	1.756	0.1771	0.0393	29	-1.03	133.99	0
133	Proline (%)	0910	1.620	0.1200	1.756	0.1771	0.0393	29	-0.77	133.00	0
133	Proline (%)	2059	1.669	0.0010	1.756	0.1771	0.0393	29	-0.50	133.00	0
133	Proline (%)	0870	1.679	0.0245	1.756	0.1771	0.0393	29	-0.44	133.00	0
133	Proline (%)	0684	1.685	0.0500	1.756	0.1771	0.0393	29	-0.40	133.00	0
133	Proline (%)	0652	1.690	0.0800	1.756	0.1771	0.0393	29	-0.38	133.00	0
133	Proline (%)	2188	1.690	0.0000	1.756	0.1771	0.0393	29	-0.38	133.05	0
133	Proline (%)	0504	1.700	0.0200	1.756	0.1771	0.0393	29	-0.32	133.00	0
133	Proline (%)	0571	1.711	0.0340	1.756	0.1771	0.0393	29	-0.26	133.00	0
133	Proline (%)	0610	1.715	0.0100	1.756	0.1771	0.0393	29	-0.23	133.05	0
133	Proline (%)	0626	1.729	0.0310	1.756	0.1771	0.0393	29	-0.16	133.05	0
133	Proline (%)	0644	1.730	0.0020	1.756	0.1771	0.0393	29	-0.15	133.00	0
133	Proline (%)	0354	1.733	0.0150	1.756	0.1771	0.0393	29	-0.14	133.00	0
133	Proline (%)	0353	1.750	0.0400	1.756	0.1771	0.0393	29	-0.04	133.00	0
133	Proline (%)	0872	1.760	0.0040	1.756	0.1771	0.0393	29	0.02	133.00	0
133	Proline (%)	0968	1.770	0.0030	1.756	0.1771	0.0393	29	0.07	133.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
133	Proline (%)	0619	1.790	0.1400	1.756	0.1771	0.0393	29	0.19	133.00	0
133	Proline (%)	0868	1.804	0.0050	1.756	0.1771	0.0393	29	0.27	133.00	0
133	Proline (%)	0227	1.815	0.1300	1.756	0.1771	0.0393	29	0.33	133.00	0
133	Proline (%)	0676	1.835	0.0900	1.756	0.1771	0.0393	29	0.44	133.05	0
133	Proline (%)	0148	1.955	0.0700	1.756	0.1771	0.0393	29	1.12	133.05	0
133	Proline (%)	0939	1.975	0.0300	1.756	0.1771	0.0393	29	1.23	133.00	0
133	Proline (%)	0918	2.025	0.0016	1.756	0.1771	0.0393	29	1.52	133.00	0
133	Proline (%)	0859	2.127	0.0604	1.756	0.1771	0.0393	29	2.09	133.00	0
133	Proline (%)	0407	2.201	0.1330	1.756	0.1771	0.0393	29	2.51	133.05	0
133	Proline (%)	2246	3.270	0.0200	1.756	0.1771	0.0393	29	8.55	133.05	0
133	Proline (%)	0226	1.837	0.3305	1.756	0.1771	0.0393	29	0.46	133.00	1
134	Serine (%)	0889	1.258	0.0046	1.532	0.1023	0.0373	30	-2.68	134.99	0
134	Serine (%)	0675	1.310	0.0200	1.532	0.1023	0.0373	30	-2.17	134.00	0
134	Serine (%)	0504	1.355	0.0100	1.532	0.1023	0.0373	30	-1.73	134.00	0
134	Serine (%)	0098	1.370	0.0660	1.532	0.1023	0.0373	30	-1.58	134.02	0
134	Serine (%)	2246	1.440	0.0800	1.532	0.1023	0.0373	30	-0.90	134.05	0
134	Serine (%)	0644	1.471	0.0070	1.532	0.1023	0.0373	30	-0.60	134.00	0
134	Serine (%)	0610	1.475	0.0300	1.532	0.1023	0.0373	30	-0.56	134.05	0
134	Serine (%)	2059	1.477	0.0190	1.532	0.1023	0.0373	30	-0.54	134.00	0
134	Serine (%)	0872	1.492	0.0080	1.532	0.1023	0.0373	30	-0.39	134.00	0
134	Serine (%)	0226	1.509	0.1320	1.532	0.1023	0.0373	30	-0.22	134.00	0
134	Serine (%)	0571	1.510	0.0260	1.532	0.1023	0.0373	30	-0.21	134.00	0
134	Serine (%)	0910	1.510	0.0800	1.532	0.1023	0.0373	30	-0.21	134.00	0
134	Serine (%)	0353	1.515	0.0500	1.532	0.1023	0.0373	30	-0.16	134.00	0
134	Serine (%)	0652	1.520	0.0000	1.532	0.1023	0.0373	30	-0.12	134.00	0
134	Serine (%)	0918	1.524	0.0204	1.532	0.1023	0.0373	30	-0.07	134.00	0
134	Serine (%)	2196	1.531	0.0000	1.532	0.1023	0.0373	30	-0.01	134.05	0
134	Serine (%)	0868	1.535	0.0120	1.532	0.1023	0.0373	30	0.03	134.00	0
134	Serine (%)	0227	1.540	0.0200	1.532	0.1023	0.0373	30	0.08	134.00	0
134	Serine (%)	0676	1.540	0.0800	1.532	0.1023	0.0373	30	0.08	134.05	0
134	Serine (%)	0859	1.549	0.0166	1.532	0.1023	0.0373	30	0.16	134.00	0
134	Serine (%)	2188	1.550	0.0000	1.532	0.1023	0.0373	30	0.18	134.05	0
134	Serine (%)	0626	1.570	0.0410	1.532	0.1023	0.0373	30	0.37	134.05	0
134	Serine (%)	0968	1.582	0.0300	1.532	0.1023	0.0373	30	0.49	134.00	0
134	Serine (%)	0354	1.613	0.0010	1.532	0.1023	0.0373	30	0.79	134.00	0
134	Serine (%)	0684	1.615	0.1500	1.532	0.1023	0.0373	30	0.81	134.00	0
134	Serine (%)	0939	1.635	0.0500	1.532	0.1023	0.0373	30	1.01	134.00	0
134	Serine (%)	0870	1.692	0.0152	1.532	0.1023	0.0373	30	1.56	134.00	0
134	Serine (%)	0682	1.711	0.0000	1.532	0.1023	0.0373	30	1.75	134.00	0
134	Serine (%)	0619	1.750	0.0600	1.532	0.1023	0.0373	30	2.13	134.00	0
134	Serine (%)	0148	1.755	0.0900	1.532	0.1023	0.0373	30	2.18	134.05	0
134	Serine (%)	0407	1.566	0.3350	1.532	0.1023	0.0373	30	0.33	134.05	1

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
135	Threonine (%)	0675	0.9100	0.0200	1.103	0.0558	0.0180	29	-3.46	135.00	0
135	Threonine (%)	0889	0.9578	0.0091	1.103	0.0558	0.0180	29	-2.61	135.99	0
135	Threonine (%)	2246	1.020	0.0600	1.103	0.0558	0.0180	29	-1.49	135.05	0
135	Threonine (%)	0226	1.041	0.0933	1.103	0.0558	0.0180	29	-1.11	135.00	0
135	Threonine (%)	2196	1.062	0.0000	1.103	0.0558	0.0180	29	-0.74	135.05	0
135	Threonine (%)	2059	1.065	0.0090	1.103	0.0558	0.0180	29	-0.69	135.00	0
135	Threonine (%)	0353	1.065	0.0100	1.103	0.0558	0.0180	29	-0.68	135.00	0
135	Threonine (%)	0652	1.070	0.0000	1.103	0.0558	0.0180	29	-0.59	135.00	0
135	Threonine (%)	0626	1.075	0.0280	1.103	0.0558	0.0180	29	-0.50	135.05	0
135	Threonine (%)	0939	1.090	0.0000	1.103	0.0558	0.0180	29	-0.24	135.00	0
135	Threonine (%)	0872	1.092	0.0290	1.103	0.0558	0.0180	29	-0.21	135.00	0
135	Threonine (%)	0682	1.095	0.0000	1.103	0.0558	0.0180	29	-0.15	135.00	0
135	Threonine (%)	0644	1.097	0.0070	1.103	0.0558	0.0180	29	-0.12	135.00	0
135	Threonine (%)	0868	1.098	0.0280	1.103	0.0558	0.0180	29	-0.09	135.00	0
135	Threonine (%)	0571	1.100	0.0310	1.103	0.0558	0.0180	29	-0.07	135.00	0
135	Threonine (%)	2188	1.100	0.0000	1.103	0.0558	0.0180	29	-0.06	135.05	0
135	Threonine (%)	0610	1.105	0.0300	1.103	0.0558	0.0180	29	0.03	135.05	0
135	Threonine (%)	0098	1.107	0.0570	1.103	0.0558	0.0180	29	0.06	135.02	0
135	Threonine (%)	0504	1.110	0.0000	1.103	0.0558	0.0180	29	0.12	135.00	0
135	Threonine (%)	0676	1.110	0.0200	1.103	0.0558	0.0180	29	0.12	135.05	0
135	Threonine (%)	0227	1.120	0.0000	1.103	0.0558	0.0180	29	0.30	135.00	0
135	Threonine (%)	0859	1.134	0.0209	1.103	0.0558	0.0180	29	0.56	135.00	0
135	Threonine (%)	0354	1.141	0.0010	1.103	0.0558	0.0180	29	0.67	135.00	0
135	Threonine (%)	0684	1.155	0.0100	1.103	0.0558	0.0180	29	0.93	135.00	0
135	Threonine (%)	0968	1.160	0.0160	1.103	0.0558	0.0180	29	1.02	135.00	0
135	Threonine (%)	0870	1.183	0.0044	1.103	0.0558	0.0180	29	1.43	135.00	0
135	Threonine (%)	0619	1.190	0.0000	1.103	0.0558	0.0180	29	1.56	135.00	0
135	Threonine (%)	0918	1.234	0.0076	1.103	0.0558	0.0180	29	2.34	135.00	0
135	Threonine (%)	0148	1.255	0.0300	1.103	0.0558	0.0180	29	2.72	135.05	0
135	Threonine (%)	0407	0.9185	0.2050	1.103	0.0558	0.0180	29	-3.31	135.05	1
135	Threonine (%)	0910	1.140	0.2000	1.103	0.0558	0.0180	29	0.66	135.00	1
136	Tryptophan (%)	0918	0.3200	0.0200	0.3995	0.0546	0.0078	19	-1.46	136.00	0
136	Tryptophan (%)	0910	0.3350	0.0300	0.3995	0.0546	0.0078	19	-1.18	136.00	0
136	Tryptophan (%)	0870	0.3401	0.0020	0.3995	0.0546	0.0078	19	-1.09	136.00	0
136	Tryptophan (%)	0682	0.3540	0.0000	0.3995	0.0546	0.0078	19	-0.83	136.00	0
136	Tryptophan (%)	0504	0.3550	0.0300	0.3995	0.0546	0.0078	19	-0.82	136.99	0
136	Tryptophan (%)	0098	0.3730	0.0020	0.3995	0.0546	0.0078	19	-0.49	136.02	0
136	Tryptophan (%)	0939	0.3900	0.0000	0.3995	0.0546	0.0078	19	-0.17	136.99	0
136	Tryptophan (%)	0859	0.3910	0.0087	0.3995	0.0546	0.0078	19	-0.16	136.03	0
136	Tryptophan (%)	0872	0.3925	0.0030	0.3995	0.0546	0.0078	19	-0.13	136.03	0
136	Tryptophan (%)	0868	0.3950	0.0020	0.3995	0.0546	0.0078	19	-0.08	136.01	0
136	Tryptophan (%)	2059	0.3995	0.0010	0.3995	0.0546	0.0078	19	0.00	136.03	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
136	Tryptophan (%)	0644	0.4045	0.0010	0.3995	0.0546	0.0078	19	0.09	136.01	0
136	Tryptophan (%)	0619	0.4115	0.0090	0.3995	0.0546	0.0078	19	0.22	136.00	0
136	Tryptophan (%)	0571	0.4115	0.0050	0.3995	0.0546	0.0078	19	0.22	136.03	0
136	Tryptophan (%)	0227	0.4400	0.0000	0.3995	0.0546	0.0078	19	0.74	136.00	0
136	Tryptophan (%)	0684	0.4400	0.0000	0.3995	0.0546	0.0078	19	0.74	136.00	0
136	Tryptophan (%)	0353	0.4750	0.0100	0.3995	0.0546	0.0078	19	1.38	136.03	0
136	Tryptophan (%)	2246	0.5700	0.0200	0.3995	0.0546	0.0078	19	3.12	136.05	0
136	Tryptophan (%)	0889	0.6439	0.0045	0.3995	0.0546	0.0078	19	4.48	136.99	0
136	Tryptophan (%)	0407	0.2830	0.1280	0.3995	0.0546	0.0078	19	-2.13	136.05	1
137	Tyrosine (%)	0889	0.6439	0.0045	1.009	0.1509	0.0285	25	-2.42	137.99	0
137	Tyrosine (%)	2246	0.7600	0.0800	1.009	0.1509	0.0285	25	-1.65	137.05	0
137	Tyrosine (%)	0675	0.8300	0.0200	1.009	0.1509	0.0285	25	-1.19	137.00	0
137	Tyrosine (%)	0968	0.8345	0.0330	1.009	0.1509	0.0285	25	-1.16	137.00	0
137	Tyrosine (%)	0626	0.8870	0.0220	1.009	0.1509	0.0285	25	-0.81	137.05	0
137	Tyrosine (%)	0910	0.9000	0.1200	1.009	0.1509	0.0285	25	-0.73	137.00	0
137	Tyrosine (%)	0870	0.9164	0.0215	1.009	0.1509	0.0285	25	-0.62	137.00	0
137	Tyrosine (%)	0918	0.9197	0.0063	1.009	0.1509	0.0285	25	-0.60	137.00	0
137	Tyrosine (%)	0939	0.9350	0.0100	1.009	0.1509	0.0285	25	-0.49	137.00	0
137	Tyrosine (%)	0227	0.9450	0.0100	1.009	0.1509	0.0285	25	-0.43	137.00	0
137	Tyrosine (%)	0098	0.9920	0.0260	1.009	0.1509	0.0285	25	-0.12	137.02	0
137	Tyrosine (%)	2196	0.9980	0.0000	1.009	0.1509	0.0285	25	-0.08	137.05	0
137	Tyrosine (%)	0354	1.012	0.0030	1.009	0.1509	0.0285	25	0.01	137.00	0
137	Tyrosine (%)	0644	1.059	0.0030	1.009	0.1509	0.0285	25	0.33	137.00	0
137	Tyrosine (%)	0682	1.069	0.0000	1.009	0.1509	0.0285	25	0.39	137.00	0
137	Tyrosine (%)	0226	1.076	0.1407	1.009	0.1509	0.0285	25	0.44	137.00	0
137	Tyrosine (%)	0872	1.081	0.0170	1.009	0.1509	0.0285	25	0.47	137.00	0
137	Tyrosine (%)	0353	1.085	0.0100	1.009	0.1509	0.0285	25	0.50	137.00	0
137	Tyrosine (%)	0684	1.090	0.0400	1.009	0.1509	0.0285	25	0.53	137.00	0
137	Tyrosine (%)	0504	1.110	0.0000	1.009	0.1509	0.0285	25	0.67	137.00	0
137	Tyrosine (%)	2188	1.120	0.0000	1.009	0.1509	0.0285	25	0.73	137.05	0
137	Tyrosine (%)	0610	1.140	0.0200	1.009	0.1509	0.0285	25	0.87	137.05	0
137	Tyrosine (%)	2059	1.201	0.0050	1.009	0.1509	0.0285	25	1.27	137.00	0
137	Tyrosine (%)	0676	1.265	0.0500	1.009	0.1509	0.0285	25	1.69	137.05	0
137	Tyrosine (%)	0148	1.375	0.0700	1.009	0.1509	0.0285	25	2.42	137.05	0
137	Tyrosine (%)	0407	0.8585	0.5430	1.009	0.1509	0.0285	25	-1.00	137.05	1
138	Valine (%)	0353	1.105	0.0100	1.435	0.1044	0.0314	30	-3.16	138.00	0
138	Valine (%)	0889	1.120	0.0138	1.435	0.1044	0.0314	30	-3.01	138.99	0
138	Valine (%)	0675	1.160	0.0200	1.435	0.1044	0.0314	30	-2.63	138.00	0
138	Valine (%)	0226	1.210	0.0242	1.435	0.1044	0.0314	30	-2.15	138.00	0
138	Valine (%)	0939	1.330	0.0400	1.435	0.1044	0.0314	30	-1.00	138.00	0
138	Valine (%)	2196	1.348	0.0000	1.435	0.1044	0.0314	30	-0.83	138.05	0
138	Valine (%)	0354	1.377	0.0190	1.435	0.1044	0.0314	30	-0.56	138.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
138	Valine (%)	0870	1.382	0.0070	1.435	0.1044	0.0314	30	-0.50	138.00	0
138	Valine (%)	0682	1.383	0.0000	1.435	0.1044	0.0314	30	-0.49	138.00	0
138	Valine (%)	0684	1.385	0.1700	1.435	0.1044	0.0314	30	-0.47	138.00	0
138	Valine (%)	0619	1.390	0.0000	1.435	0.1044	0.0314	30	-0.43	138.00	0
138	Valine (%)	0652	1.400	0.0200	1.435	0.1044	0.0314	30	-0.33	138.00	0
138	Valine (%)	0868	1.401	0.0730	1.435	0.1044	0.0314	30	-0.33	138.00	0
138	Valine (%)	0910	1.420	0.1000	1.435	0.1044	0.0314	30	-0.14	138.00	0
138	Valine (%)	2246	1.445	0.0300	1.435	0.1044	0.0314	30	0.10	138.05	0
138	Valine (%)	2059	1.448	0.0010	1.435	0.1044	0.0314	30	0.12	138.00	0
138	Valine (%)	0676	1.455	0.0300	1.435	0.1044	0.0314	30	0.20	138.05	0
138	Valine (%)	0626	1.456	0.0470	1.435	0.1044	0.0314	30	0.20	138.05	0
138	Valine (%)	0872	1.471	0.0570	1.435	0.1044	0.0314	30	0.34	138.00	0
138	Valine (%)	0571	1.491	0.0260	1.435	0.1044	0.0314	30	0.54	138.00	0
138	Valine (%)	0918	1.498	0.0521	1.435	0.1044	0.0314	30	0.61	138.00	0
138	Valine (%)	2188	1.500	0.0000	1.435	0.1044	0.0314	30	0.63	138.05	0
138	Valine (%)	0968	1.510	0.0060	1.435	0.1044	0.0314	30	0.72	138.00	0
138	Valine (%)	0098	1.511	0.0120	1.435	0.1044	0.0314	30	0.73	138.02	0
138	Valine (%)	0644	1.535	0.0040	1.435	0.1044	0.0314	30	0.96	138.00	0
138	Valine (%)	0610	1.535	0.0100	1.435	0.1044	0.0314	30	0.96	138.05	0
138	Valine (%)	0227	1.550	0.1000	1.435	0.1044	0.0314	30	1.11	138.00	0
138	Valine (%)	0859	1.554	0.0112	1.435	0.1044	0.0314	30	1.15	138.00	0
138	Valine (%)	0504	1.560	0.0400	1.435	0.1044	0.0314	30	1.20	138.00	0
138	Valine (%)	0148	1.710	0.0200	1.435	0.1044	0.0314	30	2.64	138.05	0
138	Valine (%)	0407	1.748	0.3110	1.435	0.1044	0.0314	30	3.00	138.05	1
139	Taurine (%)	0682	0.0540	0.0000				3		139.00	0
139	Taurine (%)	0889	0.0900	0.0000				3		139.99	0
139	Taurine (%)	0504	0.1300	0.0200				3		139.00	0
139	Taurine (%)	0407	0.0000	0.0000				3		139.05	4
139	Taurine (%)	0098	< 0.01					3		139.02	5
162	Glucose (%)	0227	< 0.15					0		162.99	5
163	Lactose (%)	0227	< 0.15					0		163.99	5
164	Maltose (%)	0227	< 0.15					0		164.99	5
165	Sucrose (%)	0227	3.600	0.3000				1		165.99	0
166	Raffinose (%)	0227	1.035	0.0100				1		166.99	0
167	Stachyose (%)	0227	1.385	0.0100				1		167.99	0
373	Oxytetracycline (mg/kg (ppm))	0939	0.7655	0.0150				2		373.06	0
373	Oxytetracycline (mg/kg (ppm))	2248	1.990	0.4600				2		373.06	0
373	Oxytetracycline (mg/kg (ppm))	0218	< 10					2		373.03	5
393	Ractopamine Hydrochloride (mg/kg (ppm))	2248	0.0185	0.0010				2		393.02	0
393	Ractopamine Hydrochloride (mg/kg (ppm))	2246	5.585	0.1100				2		393.99	0
393	Ractopamine Hydrochloride (mg/kg (ppm))	0148	< 0.033					2		393.02	5
393	Ractopamine Hydrochloride (mg/kg (ppm))	0904	< 1					2		393.00	5

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
400	Water Activity (Units)	2268	0.4080	0.0040	0.5108	0.0337	0.0028	10	-3.05	400.01	0
400	Water Activity (Units)	0942	0.4980	0.0060	0.5108	0.0337	0.0028	10	-0.38	400.01	0
400	Water Activity (Units)	2033	0.4985	0.0010	0.5108	0.0337	0.0028	10	-0.37	400.99	0
400	Water Activity (Units)	0407	0.5015	0.0015	0.5108	0.0337	0.0028	10	-0.28	400.01	0
400	Water Activity (Units)	0589	0.5035	0.0050	0.5108	0.0337	0.0028	10	-0.22	400.01	0
400	Water Activity (Units)	0843	0.5035	0.0030	0.5108	0.0337	0.0028	10	-0.22	400.99	0
400	Water Activity (Units)	0843	0.5078	0.0001	0.5108	0.0337	0.0028	10	-0.09	400.01	0
400	Water Activity (Units)	2075	0.5151	0.0005	0.5108	0.0337	0.0028	10	0.12	400.01	0
400	Water Activity (Units)	2109	0.5590	0.0042	0.5108	0.0337	0.0028	10	1.43	400.01	0
400	Water Activity (Units)	2073	0.5875	0.0030	0.5108	0.0337	0.0028	10	2.27	400.99	0
412	Starch, Dietary (%)	0353	10.50	0.0700				1		412.01	0
516	Arsenic, Total (mg / kg (ppm))	0186	0.1019	0.0052	0.1284	0.0196	0.0168	8	-1.35	516.52	0
516	Arsenic, Total (mg / kg (ppm))	0045	0.1130	0.0220	0.1284	0.0196	0.0168	8	-0.79	516.00	0
516	Arsenic, Total (mg / kg (ppm))	0910	0.1200	0.0000	0.1284	0.0196	0.0168	8	-0.43	516.52	0
516	Arsenic, Total (mg / kg (ppm))	0199	0.1250	0.0100	0.1284	0.0196	0.0168	8	-0.17	516.53	0
516	Arsenic, Total (mg / kg (ppm))	0171	0.1295	0.0571	0.1284	0.0196	0.0168	8	0.05	516.00	0
516	Arsenic, Total (mg / kg (ppm))	0553	0.1370	0.0180	0.1284	0.0196	0.0168	8	0.44	516.53	0
516	Arsenic, Total (mg / kg (ppm))	0918	0.1470	0.0183	0.1284	0.0196	0.0168	8	0.94	516.53	0
516	Arsenic, Total (mg / kg (ppm))	2012	0.1540	0.0040	0.1284	0.0196	0.0168	8	1.30	516.53	0
516	Arsenic, Total (mg / kg (ppm))	0407	1.166	0.6370	0.1284	0.0196	0.0168	8	52.85	516.43	2
518	Cadmium (mg / kg (ppm))	0407	0.1345	0.0113	0.1563	0.0175	0.0080	11	-1.24	518.43	0
518	Cadmium (mg / kg (ppm))	0610	0.1400	0.0200	0.1563	0.0175	0.0080	11	-0.93	518.53	0
518	Cadmium (mg / kg (ppm))	0407	0.1448	0.0025	0.1563	0.0175	0.0080	11	-0.66	518.41	0
518	Cadmium (mg / kg (ppm))	2012	0.1495	0.0010	0.1563	0.0175	0.0080	11	-0.39	518.53	0
518	Cadmium (mg / kg (ppm))	0199	0.1500	0.0000	0.1563	0.0175	0.0080	11	-0.36	518.53	0
518	Cadmium (mg / kg (ppm))	0910	0.1500	0.0200	0.1563	0.0175	0.0080	11	-0.36	518.52	0
518	Cadmium (mg / kg (ppm))	0186	0.1550	0.0080	0.1563	0.0175	0.0080	11	-0.07	518.52	0
518	Cadmium (mg / kg (ppm))	0968	0.1650	0.0020	0.1563	0.0175	0.0080	11	0.50	518.43	0
518	Cadmium (mg / kg (ppm))	0918	0.1721	0.0028	0.1563	0.0175	0.0080	11	0.90	518.53	0
518	Cadmium (mg / kg (ppm))	0553	0.1755	0.0070	0.1563	0.0175	0.0080	11	1.10	518.53	0
518	Cadmium (mg / kg (ppm))	0045	0.4055	0.0130	0.1563	0.0175	0.0080	11	14.23	518.42	0
518	Cadmium (mg / kg (ppm))	0171	0.2050	0.1300	0.1563	0.0175	0.0080	11	2.78	518.41	1
520	Chromium (mg / kg (ppm))	0186	1.935	0.0900	3.167	0.9361	0.1608	11	-1.32	520.52	0
520	Chromium (mg / kg (ppm))	0968	2.246	0.0850	3.167	0.9361	0.1608	11	-0.98	520.43	0
520	Chromium (mg / kg (ppm))	0918	2.382	0.2109	3.167	0.9361	0.1608	11	-0.84	520.53	0
520	Chromium (mg / kg (ppm))	0199	2.710	0.1800	3.167	0.9361	0.1608	11	-0.49	520.53	0
520	Chromium (mg / kg (ppm))	0407	2.872	0.3290	3.167	0.9361	0.1608	11	-0.32	520.41	0
520	Chromium (mg / kg (ppm))	0171	3.200	0.0200	3.167	0.9361	0.1608	11	0.03	520.41	0
520	Chromium (mg / kg (ppm))	0164	3.325	0.1700	3.167	0.9361	0.1608	11	0.17	520.53	0
520	Chromium (mg / kg (ppm))	0553	3.685	0.3700	3.167	0.9361	0.1608	11	0.55	520.53	0
520	Chromium (mg / kg (ppm))	0693	3.752	0.0440	3.167	0.9361	0.1608	11	0.62	520.42	0
520	Chromium (mg / kg (ppm))	0045	4.280	0.1600	3.167	0.9361	0.1608	11	1.19	520.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
520	Chromium (mg / kg (ppm))	0510	4.455	0.1100	3.167	0.9361	0.1608	11	1.38	520.43	0
520	Chromium (mg / kg (ppm))	0407	3.461	1.528	3.167	0.9361	0.1608	11	0.31	520.43	1
526	Lead (mg / kg (ppm))	0199	0.0750	0.0100	0.1276	0.0707	0.0152	8	-0.74	526.53	0
526	Lead (mg / kg (ppm))	0186	0.0758	0.0013	0.1276	0.0707	0.0152	8	-0.73	526.52	0
526	Lead (mg / kg (ppm))	0918	0.0877	0.0002	0.1276	0.0707	0.0152	8	-0.56	526.53	0
526	Lead (mg / kg (ppm))	0553	0.0932	0.0157	0.1276	0.0707	0.0152	8	-0.49	526.53	0
526	Lead (mg / kg (ppm))	2012	0.0970	0.0060	0.1276	0.0707	0.0152	8	-0.43	526.53	0
526	Lead (mg / kg (ppm))	0407	0.1525	0.0377	0.1276	0.0707	0.0152	8	0.35	526.41	0
526	Lead (mg / kg (ppm))	0407	0.2062	0.0403	0.1276	0.0707	0.0152	8	1.11	526.43	0
526	Lead (mg / kg (ppm))	0171	0.4850	0.0100	0.1276	0.0707	0.0152	8	5.06	526.41	0
526	Lead (mg / kg (ppm))	0910	< 0.1		0.1276	0.0707	0.0152	8		526.52	5
529	Mercury (µg / kg (ppb))	2012	2.109	0.2880				1		529.99	0
529	Mercury (µg / kg (ppb))	0918	< 20					1		529.99	5
529	Mercury (µg / kg (ppb))	0199	< 50					1		529.99	5
539	Nickel (mg / kg (ppm))	0186	3.175	0.0900	3.375	0.1688	0.2099	6	-1.18	539.52	0
539	Nickel (mg / kg (ppm))	0407	3.312	0.5655	3.375	0.1688	0.2099	6	-0.38	539.43	0
539	Nickel (mg / kg (ppm))	0918	3.337	0.2178	3.375	0.1688	0.2099	6	-0.23	539.53	0
539	Nickel (mg / kg (ppm))	0407	3.399	0.2658	3.375	0.1688	0.2099	6	0.14	539.41	0
539	Nickel (mg / kg (ppm))	0171	3.400	0.0800	3.375	0.1688	0.2099	6	0.15	539.41	0
539	Nickel (mg / kg (ppm))	0553	3.920	0.0400	3.375	0.1688	0.2099	6	3.23	539.53	0
702	Butyric Acid (4:0) (%)	0619	0.0000	0.0000				0		702.00	4
704	Caproic Acid (6:0) (%)	0619	0.0000	0.0000				0		704.00	4
706	Caprylic acid (8:0) (%) (w/w)	0619	0.0000	0.0000				0		706.01	4
708	Capric acid (10:0) (%) (w/w)	0619	0.0000	0.0000				0		708.01	4
710	Lauric Acid (12:0) (%) (w/w)	0619	0.0195	0.0010				1		710.01	0
710	Lauric Acid (12:0) (%) (w/w)	0226	< 0.005					1		710.99	5
710	Lauric Acid (12:0) (%) (w/w)	0676	< 0.01					1		710.99	5
710	Lauric Acid (12:0) (%) (w/w)	0164	< 0.02					1		710.99	5
714	Myristic Acid (14:0) (%) (w/w)	0226	0.0245	0.0010				3		714.99	0
714	Myristic Acid (14:0) (%) (w/w)	0619	0.0335	0.0010				3		714.01	0
714	Myristic Acid (14:0) (%) (w/w)	0164	0.0350	0.0100				3		714.99	0
716	Palmitic Acid (16:0) (%) (w/w)	0619	0.8675	0.0050				3		716.01	0
716	Palmitic Acid (16:0) (%) (w/w)	0226	0.8685	0.0210				3		716.99	0
716	Palmitic Acid (16:0) (%) (w/w)	0164	1.070	0.0600				3		716.99	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0226	0.0460	0.0020	0.0586	0.0122	0.0038	4		718.99	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0676	0.0540	0.0020	0.0586	0.0122	0.0038	4		718.99	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0619	0.0595	0.0010	0.0586	0.0122	0.0038	4		718.01	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0164	0.0750	0.0100	0.0586	0.0122	0.0038	4		718.99	0
722	Stearic Acid (18:0) (%) (w/w)	0619	0.1575	0.0030				3		722.01	0
722	Stearic Acid (18:0) (%) (w/w)	0226	0.1760	0.0040				3		722.99	0
722	Stearic Acid (18:0) (%) (w/w)	0164	0.2100	0.0200				3		722.99	0
724	Oleic Acid (9c-18:1) (%) (w/w)	0619	1.205	0.0700				3		724.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
724	Oleic Acid (9c-18:1) (% (w/w))	0226	1.447	0.0440				3	724.99	0	
724	Oleic Acid (9c-18:1) (% (w/w))	0164	1.765	0.0500				3	724.99	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0619	1.675	0.0300	2.032	0.2645	0.0178	5	726.01	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0226	1.898	0.0150	2.032	0.2645	0.0178	5	726.99	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0148	2.025	0.0300	2.032	0.2645	0.0178	5	726.02	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0164	2.205	0.0100	2.032	0.2645	0.0178	5	726.99	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0676	2.355	0.0040	2.032	0.2645	0.0178	5	726.99	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/w))	0226	0.1165	0.0010	0.1303	0.0146	0.0010	4	728.99	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/w))	0619	0.1190	0.0000	0.1303	0.0146	0.0010	4	728.01	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/w))	0164	0.1400	0.0000	0.1303	0.0146	0.0010	4	728.99	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/w))	0676	0.1455	0.0030	0.1303	0.0146	0.0010	4	728.99	0	
730	Arachidic Acid (20:0) (% (w/w))	0226	0.0140	0.0000				2	730.99	0	
730	Arachidic Acid (20:0) (% (w/w))	0619	0.0350	0.0020				2	730.01	0	
730	Arachidic Acid (20:0) (% (w/w))	0164	< 0.02					2	730.99	5	
732	Gondoic Acid (11c-20:1) (% (w/w))	0226	0.0390	0.0000				3	732.99	0	
732	Gondoic Acid (11c-20:1) (% (w/w))	0619	0.0470	0.0000				3	732.01	0	
732	Gondoic Acid (11c-20:1) (% (w/w))	0164	0.0500	0.0000				3	732.99	0	
736	Arachidonic Acid (5c,8c,11c,14c-20:4) (% (w/w))	0676	0.0100	0.0020				1	736.99	0	
736	Arachidonic Acid (5c,8c,11c,14c-20:4) (% (w/w))	0619	0.0000	0.0000				1	736.01	4	
738	Mead Acid (11c,14c,17c-20:3) (% (w/w))	0619	0.0325	0.0010				1	738.01	0	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c,20:5) (% (w/w))	0226	0.0050	0.0000				2	740.99	0	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c,20:5) (% (w/w))	0676	0.0055	0.0010				2	740.99	0	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c,20:5) (% (w/w))	0619	0.0000	0.0000				2	740.01	4	
742	Behenic Acid (22:0) (% (w/w))	0226	0.0215	0.0070				2	742.99	0	
742	Behenic Acid (22:0) (% (w/w))	0164	0.0250	0.0100				2	742.99	0	
744	Erucic Acid (13c-22:1) (% (w/w))	0619	0.0000	0.0000				0	744.01	4	
744	Erucic Acid (13c-22:1) (% (w/w))	0226	< 0.005					0	744.99	5	
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c,22:5) (% (w/w))	0676	0.0055	0.0010				1	746.99	0	
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c,22:5) (% (w/w))	0619	0.0000	0.0000				1	746.01	4	
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c,22:5) (% (w/w))	0226	< 0.005					1	746.99	5	
748	Lignoceric Acid (24:0) (% (w/w))	0226	0.0115	0.0050				1	748.99	0	
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22:6) (% (w/w))	0226	0.0100	0.0020				2	750.99	0	
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22:6) (% (w/w))	0676	0.0100	0.0000				2	750.99	0	
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22:6) (% (w/w))	0619	0.0000	0.0000				2	750.01	4	
752	Nervonic Acid (24:1) isomers (% (w/w))	0619	0.0000	0.0000				0	752.01	4	
752	Nervonic Acid (24:1) isomers (% (w/w))	0226	< 0.005					0	752.99	5	
754	Total n-3 Polyunsaturated (Omega-3) Fatty Acids (% (w/w))	0148	0.1515	0.0010				3	754.02	0	
754	Total n-3 Polyunsaturated (Omega-3) Fatty Acids (% (w/w))	0164	0.1700	0.0000				3	754.99	0	
754	Total n-3 Polyunsaturated (Omega-3) Fatty Acids (% (w/w))	0676	0.1700	0.0000				3	754.99	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Acids (% (w/w))	0148	2.050	0.0200				3	756.01	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Acids (% (w/w))	0164	2.235	0.0100				3	756.99	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Acids (% (w/w))	0676	2.405	0.0100				3	756.99	0	

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
758	Total Saturated Fatty Acids (% (w/w))	0164	1.405	0.0900				1	758.99	0	
762	Total Monounsaturated Fatty Acids (% (w/w))	0164	1.880	0.0600				1	762.99	0	
766	Total Polyunsaturated Fatty Acids (% (w/w))	0164	2.420	0.0200				1	766.99	0	
770	Total Fat (equivalent to NLEA) (% (w/w))	0164	6.000	0.1800				1	770.99	0	
772	Total Fatty Acids (% (w/w))	0226	4.678	0.0890				2	772.99	0	
772	Total Fatty Acids (% (w/w))	0164	5.740	0.1800				2	772.99	0	

Notes: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2 and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = Rejected for duplicates too far apart, 2 = Rejected as outlier, 5 = A reporting limit and 4 = zeros submitted as values. Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). Flag 3 indicates not used in statistics.