

Animal Feed Scheme

Horse Feed

Test Material Code # 202022

Labs Reporting: 177

Analytes Reported 126

Issue Date : 03/31/2020

Analyte Proficiency Testing Report

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
000	Urea (%)	0278	0.3500	0.1000				1		000.02	0
001	Loss on Drying (%)	0560	0.1000	0.0040	9.396	0.2712	0.0852	77	-34.28	001.99	0
001	Loss on Drying (%)	0618	7.650	0.2600	9.396	0.2712	0.0852	77	-6.44	001.99	0
001	Loss on Drying (%)	0038	7.790	0.5600	9.396	0.2712	0.0852	77	-5.92	001.07	0
001	Loss on Drying (%)	0019	8.350	0.9000	9.396	0.2712	0.0852	77	-3.86	001.07	0
001	Loss on Drying (%)	2155	8.350	0.2800	9.396	0.2712	0.0852	77	-3.86	001.99	0
001	Loss on Drying (%)	0676	8.395	0.0700	9.396	0.2712	0.0852	77	-3.69	001.99	0
001	Loss on Drying (%)	0536	8.675	0.0500	9.396	0.2712	0.0852	77	-2.66	001.99	0
001	Loss on Drying (%)	0297	8.850	0.3000	9.396	0.2712	0.0852	77	-2.01	001.07	0
001	Loss on Drying (%)	0034	8.945	0.2300	9.396	0.2712	0.0852	77	-1.66	001.07	0
001	Loss on Drying (%)	0853	8.950	0.1000	9.396	0.2712	0.0852	77	-1.65	001.07	0
001	Loss on Drying (%)	2081	9.015	0.0300	9.396	0.2712	0.0852	77	-1.41	001.99	0
001	Loss on Drying (%)	0345	9.020	0.0200	9.396	0.2712	0.0852	77	-1.39	001.07	0
001	Loss on Drying (%)	2146	9.105	0.0100	9.396	0.2712	0.0852	77	-1.07	001.07	0
001	Loss on Drying (%)	0918	9.115	0.0900	9.396	0.2712	0.0852	77	-1.04	001.99	0
001	Loss on Drying (%)	0089	9.130	0.0000	9.396	0.2712	0.0852	77	-0.98	001.07	0
001	Loss on Drying (%)	0074	9.175	0.0500	9.396	0.2712	0.0852	77	-0.82	001.07	0
001	Loss on Drying (%)	2246	9.180	0.0800	9.396	0.2712	0.0852	77	-0.80	001.99	0
001	Loss on Drying (%)	0027	9.193	0.0330	9.396	0.2712	0.0852	77	-0.75	001.00	0
001	Loss on Drying (%)	0049	9.220	0.3400	9.396	0.2712	0.0852	77	-0.65	001.07	0
001	Loss on Drying (%)	2259	9.237	0.0410	9.396	0.2712	0.0852	77	-0.59	001.07	0
001	Loss on Drying (%)	0169	9.285	0.0300	9.396	0.2712	0.0852	77	-0.41	001.00	0
001	Loss on Drying (%)	0226	9.285	0.0100	9.396	0.2712	0.0852	77	-0.41	001.07	0
001	Loss on Drying (%)	0083	9.295	0.0300	9.396	0.2712	0.0852	77	-0.37	001.07	0
001	Loss on Drying (%)	0413	9.300	0.0000	9.396	0.2712	0.0852	77	-0.35	001.07	0
001	Loss on Drying (%)	0510	9.300	0.0000	9.396	0.2712	0.0852	77	-0.35	001.99	0
001	Loss on Drying (%)	2144	9.305	0.0700	9.396	0.2712	0.0852	77	-0.34	001.07	0
001	Loss on Drying (%)	0505	9.310	0.1200	9.396	0.2712	0.0852	77	-0.32	001.07	0
001	Loss on Drying (%)	0948	9.344	0.0250	9.396	0.2712	0.0852	77	-0.19	001.99	0
001	Loss on Drying (%)	0353	9.350	0.1000	9.396	0.2712	0.0852	77	-0.17	001.07	0
001	Loss on Drying (%)	0610	9.350	0.1000	9.396	0.2712	0.0852	77	-0.17	001.99	0
001	Loss on Drying (%)	0309	9.380	0.0400	9.396	0.2712	0.0852	77	-0.06	001.00	0
001	Loss on Drying (%)	0037	9.380	0.0200	9.396	0.2712	0.0852	77	-0.06	001.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
001	Loss on Drying (%)	2295	9.385	0.0100	9.396	0.2712	0.0852	77	-0.04	001.03	0
001	Loss on Drying (%)	0278	9.390	0.1200	9.396	0.2712	0.0852	77	-0.02	001.07	0
001	Loss on Drying (%)	0910	9.400	0.2000	9.396	0.2712	0.0852	77	0.01	001.99	0
001	Loss on Drying (%)	2059	9.405	0.0100	9.396	0.2712	0.0852	77	0.03	001.07	0
001	Loss on Drying (%)	0035	9.415	0.0300	9.396	0.2712	0.0852	77	0.07	001.07	0
001	Loss on Drying (%)	0139	9.425	0.0100	9.396	0.2712	0.0852	77	0.11	001.07	0
001	Loss on Drying (%)	2009	9.426	0.0138	9.396	0.2712	0.0852	77	0.11	001.07	0
001	Loss on Drying (%)	0512	9.435	0.2840	9.396	0.2712	0.0852	77	0.14	001.07	0
001	Loss on Drying (%)	0675	9.435	0.0100	9.396	0.2712	0.0852	77	0.14	001.07	0
001	Loss on Drying (%)	0066	9.440	0.1800	9.396	0.2712	0.0852	77	0.16	001.07	0
001	Loss on Drying (%)	0171	9.455	0.0100	9.396	0.2712	0.0852	77	0.22	001.07	0
001	Loss on Drying (%)	0571	9.485	0.0500	9.396	0.2712	0.0852	77	0.33	001.07	0
001	Loss on Drying (%)	2109	9.485	0.1300	9.396	0.2712	0.0852	77	0.33	001.07	0
001	Loss on Drying (%)	0199	9.495	0.0700	9.396	0.2712	0.0852	77	0.36	001.07	0
001	Loss on Drying (%)	2053	9.495	0.1500	9.396	0.2712	0.0852	77	0.36	001.07	0
001	Loss on Drying (%)	0407	9.499	0.0420	9.396	0.2712	0.0852	77	0.38	001.07	0
001	Loss on Drying (%)	0148	9.510	0.0200	9.396	0.2712	0.0852	77	0.42	001.07	0
001	Loss on Drying (%)	0590	9.510	0.2200	9.396	0.2712	0.0852	77	0.42	001.07	0
001	Loss on Drying (%)	0843	9.515	0.0300	9.396	0.2712	0.0852	77	0.44	001.03	0
001	Loss on Drying (%)	0843	9.515	0.0300	9.396	0.2712	0.0852	77	0.44	001.07	0
001	Loss on Drying (%)	2062	9.532	0.0248	9.396	0.2712	0.0852	77	0.50	001.03	0
001	Loss on Drying (%)	0868	9.545	0.0100	9.396	0.2712	0.0852	77	0.55	001.03	0
001	Loss on Drying (%)	2012	9.545	0.0500	9.396	0.2712	0.0852	77	0.55	001.07	0
001	Loss on Drying (%)	0033	9.550	0.0200	9.396	0.2712	0.0852	77	0.57	001.07	0
001	Loss on Drying (%)	0003	9.580	0.0200	9.396	0.2712	0.0852	77	0.68	001.07	0
001	Loss on Drying (%)	2192	9.585	0.0100	9.396	0.2712	0.0852	77	0.70	001.07	0
001	Loss on Drying (%)	0878	9.590	0.0200	9.396	0.2712	0.0852	77	0.71	001.03	0
001	Loss on Drying (%)	0644	9.595	0.0100	9.396	0.2712	0.0852	77	0.73	001.05	0
001	Loss on Drying (%)	0366	9.600	0.0000	9.396	0.2712	0.0852	77	0.75	001.07	0
001	Loss on Drying (%)	0689	9.600	0.0000	9.396	0.2712	0.0852	77	0.75	001.07	0
001	Loss on Drying (%)	0912	9.600	0.2000	9.396	0.2712	0.0852	77	0.75	001.99	0
001	Loss on Drying (%)	0142	9.605	0.0500	9.396	0.2712	0.0852	77	0.77	001.07	0
001	Loss on Drying (%)	0619	9.615	0.0100	9.396	0.2712	0.0852	77	0.81	001.03	0
001	Loss on Drying (%)	0028	9.620	0.0200	9.396	0.2712	0.0852	77	0.83	001.07	0
001	Loss on Drying (%)	0693	9.637	0.0834	9.396	0.2712	0.0852	77	0.89	001.07	0
001	Loss on Drying (%)	0893	9.645	0.0100	9.396	0.2712	0.0852	77	0.92	001.03	0
001	Loss on Drying (%)	0581	9.655	0.0300	9.396	0.2712	0.0852	77	0.95	001.07	0
001	Loss on Drying (%)	0876	9.700	0.2000	9.396	0.2712	0.0852	77	1.12	001.07	0
001	Loss on Drying (%)	0019	9.710	0.0200	9.396	0.2712	0.0852	77	1.16	001.99	0
001	Loss on Drying (%)	0098	9.765	0.0700	9.396	0.2712	0.0852	77	1.36	001.07	0
001	Loss on Drying (%)	0852	9.765	0.0100	9.396	0.2712	0.0852	77	1.36	001.99	0
001	Loss on Drying (%)	0872	9.775	0.0300	9.396	0.2712	0.0852	77	1.40	001.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
001	Loss on Drying (%)	0357	9.821	0.0133	9.396	0.2712	0.0852	77	1.56	001.99	0
001	Loss on Drying (%)	0968	9.848	0.0340	9.396	0.2712	0.0852	77	1.67	001.05	0
001	Loss on Drying (%)	0682	10.54	0.0000	9.396	0.2712	0.0852	77	4.22	001.07	0
001	Loss on Drying (%)	0889	31.72	45.14	9.396	0.2712	0.0852	77	82.33	001.99	2
002	Protein, Crude (%)	0298	10.24	0.0200	12.55	0.2648	0.1327	173	-8.71	002.06	0
002	Protein, Crude (%)	0852	11.58	0.0400	12.55	0.2648	0.1327	173	-3.65	002.05	0
002	Protein, Crude (%)	0014	11.80	0.0000	12.55	0.2648	0.1327	173	-2.81	002.06	0
002	Protein, Crude (%)	0598	11.92	0.1000	12.55	0.2648	0.1327	173	-2.36	002.06	0
002	Protein, Crude (%)	0529	11.96	0.0400	12.55	0.2648	0.1327	173	-2.21	002.06	0
002	Protein, Crude (%)	0539	11.98	0.4900	12.55	0.2648	0.1327	173	-2.15	002.06	0
002	Protein, Crude (%)	2053	12.05	0.1000	12.55	0.2648	0.1327	173	-1.87	002.06	0
002	Protein, Crude (%)	0939	12.15	0.2200	12.55	0.2648	0.1327	173	-1.49	002.05	0
002	Protein, Crude (%)	0596	12.16	0.0800	12.55	0.2648	0.1327	173	-1.46	002.05	0
002	Protein, Crude (%)	2246	12.16	0.0400	12.55	0.2648	0.1327	173	-1.46	002.05	0
002	Protein, Crude (%)	2062	12.16	0.0672	12.55	0.2648	0.1327	173	-1.44	002.05	0
002	Protein, Crude (%)	0676	12.19	0.1200	12.55	0.2648	0.1327	173	-1.34	002.06	0
002	Protein, Crude (%)	0619	12.20	0.0000	12.55	0.2648	0.1327	173	-1.30	002.06	0
002	Protein, Crude (%)	0006	12.21	0.0500	12.55	0.2648	0.1327	173	-1.29	002.06	0
002	Protein, Crude (%)	0199	12.22	0.0300	12.55	0.2648	0.1327	173	-1.25	002.01	0
002	Protein, Crude (%)	0563	12.23	0.1711	12.55	0.2648	0.1327	173	-1.20	002.08	0
002	Protein, Crude (%)	0675	12.23	0.0000	12.55	0.2648	0.1327	173	-1.19	002.05	0
002	Protein, Crude (%)	2181	12.25	0.0700	12.55	0.2648	0.1327	173	-1.13	002.05	0
002	Protein, Crude (%)	0910	12.25	0.1000	12.55	0.2648	0.1327	173	-1.12	002.01	0
002	Protein, Crude (%)	0853	12.25	0.3000	12.55	0.2648	0.1327	173	-1.12	002.06	0
002	Protein, Crude (%)	0876	12.25	0.1000	12.55	0.2648	0.1327	173	-1.12	002.06	0
002	Protein, Crude (%)	2109	12.26	0.1300	12.55	0.2648	0.1327	173	-1.10	002.05	0
002	Protein, Crude (%)	0653	12.26	0.0000	12.55	0.2648	0.1327	173	-1.08	002.01	0
002	Protein, Crude (%)	0868	12.27	0.0500	12.55	0.2648	0.1327	173	-1.06	002.06	0
002	Protein, Crude (%)	0939	12.27	0.1800	12.55	0.2648	0.1327	173	-1.04	002.06	0
002	Protein, Crude (%)	2304	12.28	0.0560	12.55	0.2648	0.1327	173	-1.02	002.09	0
002	Protein, Crude (%)	0169	12.29	0.0500	12.55	0.2648	0.1327	173	-0.98	002.00	0
002	Protein, Crude (%)	0144	12.29	0.2500	12.55	0.2648	0.1327	173	-0.98	002.06	0
002	Protein, Crude (%)	0644	12.29	0.7170	12.55	0.2648	0.1327	173	-0.98	002.06	0
002	Protein, Crude (%)	0620	12.29	0.0977	12.55	0.2648	0.1327	173	-0.97	002.05	0
002	Protein, Crude (%)	0870	12.29	0.2600	12.55	0.2648	0.1327	173	-0.95	002.01	0
002	Protein, Crude (%)	0036	12.30	0.1731	12.55	0.2648	0.1327	173	-0.94	002.02	0
002	Protein, Crude (%)	0652	12.30	0.2000	12.55	0.2648	0.1327	173	-0.93	002.01	0
002	Protein, Crude (%)	0689	12.30	0.0000	12.55	0.2648	0.1327	173	-0.93	002.05	0
002	Protein, Crude (%)	0297	12.33	0.0700	12.55	0.2648	0.1327	173	-0.83	002.06	0
002	Protein, Crude (%)	0897	12.33	0.0600	12.55	0.2648	0.1327	173	-0.81	002.05	0
002	Protein, Crude (%)	2199	12.35	0.5000	12.55	0.2648	0.1327	173	-0.74	002.05	0
002	Protein, Crude (%)	0956	12.35	0.1000	12.55	0.2648	0.1327	173	-0.74	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	2192	12.36	0.0000	12.55	0.2648	0.1327	173	-0.70	002.05	0
002	Protein, Crude (%)	0868	12.37	0.0800	12.55	0.2648	0.1327	173	-0.66	002.05	0
002	Protein, Crude (%)	0881	12.38	0.0680	12.55	0.2648	0.1327	173	-0.64	002.05	0
002	Protein, Crude (%)	0354	12.38	0.0200	12.55	0.2648	0.1327	173	-0.62	002.05	0
002	Protein, Crude (%)	0893	12.39	0.0000	12.55	0.2648	0.1327	173	-0.59	002.05	0
002	Protein, Crude (%)	2012	12.40	0.0100	12.55	0.2648	0.1327	173	-0.57	002.05	0
002	Protein, Crude (%)	0619	12.40	0.0000	12.55	0.2648	0.1327	173	-0.55	002.05	0
002	Protein, Crude (%)	0033	12.40	0.2000	12.55	0.2648	0.1327	173	-0.55	002.06	0
002	Protein, Crude (%)	0175	12.40	0.0000	12.55	0.2648	0.1327	173	-0.55	002.06	0
002	Protein, Crude (%)	0650	12.40	0.2000	12.55	0.2648	0.1327	173	-0.55	002.06	0
002	Protein, Crude (%)	0066	12.41	0.0700	12.55	0.2648	0.1327	173	-0.53	002.02	0
002	Protein, Crude (%)	2168	12.41	0.0200	12.55	0.2648	0.1327	173	-0.51	002.06	0
002	Protein, Crude (%)	2022	12.42	0.0100	12.55	0.2648	0.1327	173	-0.49	002.05	0
002	Protein, Crude (%)	0004	12.42	0.2600	12.55	0.2648	0.1327	173	-0.47	002.06	0
002	Protein, Crude (%)	0042	12.42	0.2800	12.55	0.2648	0.1327	173	-0.47	002.06	0
002	Protein, Crude (%)	0872	12.42	0.4800	12.55	0.2648	0.1327	173	-0.47	002.06	0
002	Protein, Crude (%)	2146	12.44	0.0600	12.55	0.2648	0.1327	173	-0.40	002.05	0
002	Protein, Crude (%)	0968	12.44	0.3400	12.55	0.2648	0.1327	173	-0.39	002.01	0
002	Protein, Crude (%)	0675	12.45	0.0100	12.55	0.2648	0.1327	173	-0.38	002.06	0
002	Protein, Crude (%)	0878	12.45	0.1400	12.55	0.2648	0.1327	173	-0.36	002.05	0
002	Protein, Crude (%)	0098	12.45	0.3000	12.55	0.2648	0.1327	173	-0.36	002.06	0
002	Protein, Crude (%)	0953	12.45	0.3000	12.55	0.2648	0.1327	173	-0.36	002.06	0
002	Protein, Crude (%)	0019	12.46	0.1500	12.55	0.2648	0.1327	173	-0.34	002.06	0
002	Protein, Crude (%)	0100	12.46	0.1500	12.55	0.2648	0.1327	173	-0.34	002.06	0
002	Protein, Crude (%)	0242	12.46	0.2300	12.55	0.2648	0.1327	173	-0.34	002.06	0
002	Protein, Crude (%)	0032	12.46	0.1260	12.55	0.2648	0.1327	173	-0.34	002.06	0
002	Protein, Crude (%)	0098	12.46	0.0800	12.55	0.2648	0.1327	173	-0.32	002.08	0
002	Protein, Crude (%)	0685	12.47	0.0100	12.55	0.2648	0.1327	173	-0.30	002.01	0
002	Protein, Crude (%)	0354	12.47	0.0300	12.55	0.2648	0.1327	173	-0.30	002.06	0
002	Protein, Crude (%)	2009	12.47	0.0082	12.55	0.2648	0.1327	173	-0.30	002.05	0
002	Protein, Crude (%)	0164	12.47	0.0600	12.55	0.2648	0.1327	173	-0.28	002.06	0
002	Protein, Crude (%)	0353	12.47	0.0400	12.55	0.2648	0.1327	173	-0.28	002.06	0
002	Protein, Crude (%)	0038	12.48	0.3050	12.55	0.2648	0.1327	173	-0.26	002.06	0
002	Protein, Crude (%)	0914	12.48	0.0200	12.55	0.2648	0.1327	173	-0.25	002.01	0
002	Protein, Crude (%)	0674	12.48	0.0000	12.55	0.2648	0.1327	173	-0.25	002.05	0
002	Protein, Crude (%)	0948	12.48	0.0400	12.55	0.2648	0.1327	173	-0.25	002.05	0
002	Protein, Crude (%)	0918	12.48	0.0800	12.55	0.2648	0.1327	173	-0.25	002.06	0
002	Protein, Crude (%)	2073	12.49	0.0200	12.55	0.2648	0.1327	173	-0.21	002.05	0
002	Protein, Crude (%)	0425	12.49	0.0600	12.55	0.2648	0.1327	173	-0.21	002.06	0
002	Protein, Crude (%)	0263	12.49	0.0072	12.55	0.2648	0.1327	173	-0.19	002.06	0
002	Protein, Crude (%)	0202	12.50	0.0900	12.55	0.2648	0.1327	173	-0.19	002.06	0
002	Protein, Crude (%)	0848	12.50	0.0300	12.55	0.2648	0.1327	173	-0.19	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0693	12.50	0.0825	12.55	0.2648	0.1327	173	-0.18	002.06	0
002	Protein, Crude (%)	2196	12.50	0.0000	12.55	0.2648	0.1327	173	-0.17	002.01	0
002	Protein, Crude (%)	0017	12.50	0.0000	12.55	0.2648	0.1327	173	-0.17	002.06	0
002	Protein, Crude (%)	0074	12.50	0.1600	12.55	0.2648	0.1327	173	-0.17	002.06	0
002	Protein, Crude (%)	0622	12.50	0.4590	12.55	0.2648	0.1327	173	-0.17	002.05	0
002	Protein, Crude (%)	0062	12.51	0.0880	12.55	0.2648	0.1327	173	-0.15	002.06	0
002	Protein, Crude (%)	0123	12.52	0.0800	12.55	0.2648	0.1327	173	-0.10	002.06	0
002	Protein, Crude (%)	0142	12.52	0.0000	12.55	0.2648	0.1327	173	-0.10	002.06	0
002	Protein, Crude (%)	0964	12.53	0.0900	12.55	0.2648	0.1327	173	-0.08	002.06	0
002	Protein, Crude (%)	2193	12.54	0.2100	12.55	0.2648	0.1327	173	-0.04	002.01	0
002	Protein, Crude (%)	0003	12.54	0.0300	12.55	0.2648	0.1327	173	-0.04	002.06	0
002	Protein, Crude (%)	0687	12.54	0.1300	12.55	0.2648	0.1327	173	-0.04	002.06	0
002	Protein, Crude (%)	0859	12.54	0.1010	12.55	0.2648	0.1327	173	-0.03	002.06	0
002	Protein, Crude (%)	0870	12.54	0.0375	12.55	0.2648	0.1327	173	-0.03	002.06	0
002	Protein, Crude (%)	2059	12.54	0.0375	12.55	0.2648	0.1327	173	-0.01	002.06	0
002	Protein, Crude (%)	0171	12.55	0.1000	12.55	0.2648	0.1327	173	0.02	002.06	0
002	Protein, Crude (%)	0590	12.55	0.1000	12.55	0.2648	0.1327	173	0.02	002.06	0
002	Protein, Crude (%)	0610	12.55	0.1000	12.55	0.2648	0.1327	173	0.02	002.06	0
002	Protein, Crude (%)	2144	12.55	0.4400	12.55	0.2648	0.1327	173	0.02	002.06	0
002	Protein, Crude (%)	0035	12.56	0.0000	12.55	0.2648	0.1327	173	0.06	002.06	0
002	Protein, Crude (%)	0199	12.58	0.0500	12.55	0.2648	0.1327	173	0.11	002.06	0
002	Protein, Crude (%)	0505	12.58	0.2400	12.55	0.2648	0.1327	173	0.13	002.06	0
002	Protein, Crude (%)	0646	12.58	0.0200	12.55	0.2648	0.1327	173	0.13	002.06	0
002	Protein, Crude (%)	0357	12.59	0.0625	12.55	0.2648	0.1327	173	0.18	002.06	0
002	Protein, Crude (%)	0358	12.60	0.0900	12.55	0.2648	0.1327	173	0.19	002.06	0
002	Protein, Crude (%)	0049	12.60	0.2000	12.55	0.2648	0.1327	173	0.21	002.06	0
002	Protein, Crude (%)	0278	12.60	0.0000	12.55	0.2648	0.1327	173	0.21	002.06	0
002	Protein, Crude (%)	0265	12.60	0.4000	12.55	0.2648	0.1327	173	0.21	002.06	0
002	Protein, Crude (%)	2259	12.61	0.1500	12.55	0.2648	0.1327	173	0.23	002.01	0
002	Protein, Crude (%)	0027	12.61	0.1760	12.55	0.2648	0.1327	173	0.24	002.06	0
002	Protein, Crude (%)	2006	12.61	0.0600	12.55	0.2648	0.1327	173	0.24	002.05	0
002	Protein, Crude (%)	0553	12.61	0.3800	12.55	0.2648	0.1327	173	0.24	002.06	0
002	Protein, Crude (%)	0148	12.61	0.0030	12.55	0.2648	0.1327	173	0.26	002.06	0
002	Protein, Crude (%)	2213	12.62	0.0300	12.55	0.2648	0.1327	173	0.26	002.06	0
002	Protein, Crude (%)	0541	12.62	0.3200	12.55	0.2648	0.1327	173	0.28	002.06	0
002	Protein, Crude (%)	2246	12.62	0.0200	12.55	0.2648	0.1327	173	0.28	002.06	0
002	Protein, Crude (%)	0139	12.64	0.4100	12.55	0.2648	0.1327	173	0.34	002.06	0
002	Protein, Crude (%)	0910	12.65	0.1000	12.55	0.2648	0.1327	173	0.40	002.06	0
002	Protein, Crude (%)	0366	12.65	0.3000	12.55	0.2648	0.1327	173	0.40	002.06	0
002	Protein, Crude (%)	0028	12.66	0.2200	12.55	0.2648	0.1327	173	0.43	002.06	0
002	Protein, Crude (%)	0083	12.67	0.0200	12.55	0.2648	0.1327	173	0.47	002.06	0
002	Protein, Crude (%)	0037	12.69	0.0100	12.55	0.2648	0.1327	173	0.53	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0309	12.69	0.2560	12.55	0.2648	0.1327	173	0.55	002.06	0
002	Protein, Crude (%)	0009	12.70	0.2500	12.55	0.2648	0.1327	173	0.56	002.06	0
002	Protein, Crude (%)	2089	12.70	0.5100	12.55	0.2648	0.1327	173	0.56	002.06	0
002	Protein, Crude (%)	0294	12.70	0.0000	12.55	0.2648	0.1327	173	0.58	002.06	0
002	Protein, Crude (%)	0510	12.70	0.0000	12.55	0.2648	0.1327	173	0.58	002.06	0
002	Protein, Crude (%)	0508	12.70	0.2800	12.55	0.2648	0.1327	173	0.60	002.06	0
002	Protein, Crude (%)	2150	12.71	0.1700	12.55	0.2648	0.1327	173	0.60	002.06	0
002	Protein, Crude (%)	0726	12.71	0.0100	12.55	0.2648	0.1327	173	0.60	002.06	0
002	Protein, Crude (%)	2109	12.71	0.1000	12.55	0.2648	0.1327	173	0.62	002.06	0
002	Protein, Crude (%)	2012	12.72	0.0100	12.55	0.2648	0.1327	173	0.64	002.06	0
002	Protein, Crude (%)	0968	12.73	0.3610	12.55	0.2648	0.1327	173	0.68	002.06	0
002	Protein, Crude (%)	0036	12.73	0.0200	12.55	0.2648	0.1327	173	0.70	002.06	0
002	Protein, Crude (%)	0226	12.74	0.4700	12.55	0.2648	0.1327	173	0.72	002.06	0
002	Protein, Crude (%)	0512	12.74	0.1500	12.55	0.2648	0.1327	173	0.72	002.06	0
002	Protein, Crude (%)	0682	12.74	0.0000	12.55	0.2648	0.1327	173	0.73	002.06	0
002	Protein, Crude (%)	0571	12.75	0.1300	12.55	0.2648	0.1327	173	0.75	002.06	0
002	Protein, Crude (%)	2033	12.75	0.0700	12.55	0.2648	0.1327	173	0.75	002.06	0
002	Protein, Crude (%)	0536	12.75	0.1000	12.55	0.2648	0.1327	173	0.77	002.05	0
002	Protein, Crude (%)	0574	12.75	0.1000	12.55	0.2648	0.1327	173	0.77	002.06	0
002	Protein, Crude (%)	0089	12.76	0.0000	12.55	0.2648	0.1327	173	0.81	002.06	0
002	Protein, Crude (%)	0596	12.79	0.0500	12.55	0.2648	0.1327	173	0.90	002.06	0
002	Protein, Crude (%)	0010	12.80	0.2000	12.55	0.2648	0.1327	173	0.96	002.06	0
002	Protein, Crude (%)	0345	12.83	0.0500	12.55	0.2648	0.1327	173	1.06	002.06	0
002	Protein, Crude (%)	0948	12.84	0.1000	12.55	0.2648	0.1327	173	1.11	002.06	0
002	Protein, Crude (%)	0589	12.85	0.0200	12.55	0.2648	0.1327	173	1.15	002.06	0
002	Protein, Crude (%)	0884	12.85	0.1000	12.55	0.2648	0.1327	173	1.15	002.06	0
002	Protein, Crude (%)	2081	12.88	0.0800	12.55	0.2648	0.1327	173	1.26	002.06	0
002	Protein, Crude (%)	0190	12.89	0.1800	12.55	0.2648	0.1327	173	1.30	002.06	0
002	Protein, Crude (%)	0413	12.90	0.2000	12.55	0.2648	0.1327	173	1.34	002.06	0
002	Protein, Crude (%)	0001	12.92	0.1600	12.55	0.2648	0.1327	173	1.41	002.06	0
002	Protein, Crude (%)	2174	12.94	0.3500	12.55	0.2648	0.1327	173	1.47	002.06	0
002	Protein, Crude (%)	0018	12.94	0.0600	12.55	0.2648	0.1327	173	1.49	002.06	0
002	Protein, Crude (%)	0407	12.95	0.2130	12.55	0.2648	0.1327	173	1.53	002.06	0
002	Protein, Crude (%)	0229	12.96	0.1220	12.55	0.2648	0.1327	173	1.58	002.06	0
002	Protein, Crude (%)	2200	13.00	0.0000	12.55	0.2648	0.1327	173	1.72	002.05	0
002	Protein, Crude (%)	2076	13.02	0.2300	12.55	0.2648	0.1327	173	1.79	002.06	0
002	Protein, Crude (%)	0034	13.03	0.2420	12.55	0.2648	0.1327	173	1.85	002.06	0
002	Protein, Crude (%)	0692	13.10	0.2000	12.55	0.2648	0.1327	173	2.09	002.06	0
002	Protein, Crude (%)	0563	13.12	0.5970	12.55	0.2648	0.1327	173	2.16	002.06	0
002	Protein, Crude (%)	0511	13.15	0.3400	12.55	0.2648	0.1327	173	2.28	002.06	0
002	Protein, Crude (%)	0618	13.29	0.0480	12.55	0.2648	0.1327	173	2.80	002.06	0
002	Protein, Crude (%)	0843	13.32	0.2600	12.55	0.2648	0.1327	173	2.93	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0047	13.40	0.6000	12.55	0.2648	0.1327	173	3.23	002.06	0
002	Protein, Crude (%)	2155	13.57	0.1550	12.55	0.2648	0.1327	173	3.88	002.06	0
002	Protein, Crude (%)	0520	13.60	0.0000	12.55	0.2648	0.1327	173	3.98	002.06	0
002	Protein, Crude (%)	2295	13.95	0.1000	12.55	0.2648	0.1327	173	5.30	002.05	0
002	Protein, Crude (%)	0553	13.98	0.0900	12.55	0.2648	0.1327	173	5.40	002.11	0
002	Protein, Crude (%)	0889	14.02	0.0800	12.55	0.2648	0.1327	173	5.57	002.11	0
002	Protein, Crude (%)	0852	14.80	0.0500	12.55	0.2648	0.1327	173	8.50	002.11	0
002	Protein, Crude (%)	0297	14.89	0.0700	12.55	0.2648	0.1327	173	8.84	002.11	0
002	Protein, Crude (%)	0405	17.80	0.1000	12.55	0.2648	0.1327	173	19.84	002.11	0
002	Protein, Crude (%)	0889	9.795	6.330	12.55	0.2648	0.1327	173	-10.39	002.01	1
002	Protein, Crude (%)	2129	32.08	0.0320	12.55	0.2648	0.1327	173	73.76	002.06	2
002	Protein, Crude (%)	2129	32.78	0.8200	12.55	0.2648	0.1327	173	76.40	002.11	2
003	Fat, Crude (%)	0536	3.165	0.0300	3.973	0.3626	0.0936	131	-2.23	003.99	0
003	Fat, Crude (%)	0893	3.210	0.1200	3.973	0.3626	0.0936	131	-2.10	003.12	0
003	Fat, Crude (%)	0037	3.245	0.0100	3.973	0.3626	0.0936	131	-2.01	003.14	0
003	Fat, Crude (%)	0581	3.310	0.0600	3.973	0.3626	0.0936	131	-1.83	003.14	0
003	Fat, Crude (%)	0897	3.315	0.0900	3.973	0.3626	0.0936	131	-1.82	003.10	0
003	Fat, Crude (%)	0574	3.375	0.0300	3.973	0.3626	0.0936	131	-1.65	003.14	0
003	Fat, Crude (%)	0939	3.425	0.0900	3.973	0.3626	0.0936	131	-1.51	003.06	0
003	Fat, Crude (%)	0144	3.445	0.0300	3.973	0.3626	0.0936	131	-1.46	003.14	0
003	Fat, Crude (%)	0242	3.455	0.0500	3.973	0.3626	0.0936	131	-1.43	003.10	0
003	Fat, Crude (%)	0298	3.470	0.0200	3.973	0.3626	0.0936	131	-1.39	003.14	0
003	Fat, Crude (%)	0520	3.500	0.0400	3.973	0.3626	0.0936	131	-1.31	003.14	0
003	Fat, Crude (%)	0083	3.515	0.0300	3.973	0.3626	0.0936	131	-1.26	003.14	0
003	Fat, Crude (%)	0019	3.530	0.0600	3.973	0.3626	0.0936	131	-1.22	003.14	0
003	Fat, Crude (%)	0036	3.583	0.0506	3.973	0.3626	0.0936	131	-1.08	003.10	0
003	Fat, Crude (%)	2302	3.590	0.3400	3.973	0.3626	0.0936	131	-1.06	003.14	0
003	Fat, Crude (%)	2076	3.609	0.1823	3.973	0.3626	0.0936	131	-1.00	003.14	0
003	Fat, Crude (%)	0034	3.610	0.1000	3.973	0.3626	0.0936	131	-1.00	003.14	0
003	Fat, Crude (%)	0912	3.620	0.0600	3.973	0.3626	0.0936	131	-0.97	003.14	0
003	Fat, Crude (%)	0876	3.650	0.1000	3.973	0.3626	0.0936	131	-0.89	003.14	0
003	Fat, Crude (%)	0263	3.650	0.0200	3.973	0.3626	0.0936	131	-0.89	003.10	0
003	Fat, Crude (%)	0358	3.660	0.0800	3.973	0.3626	0.0936	131	-0.86	003.14	0
003	Fat, Crude (%)	0505	3.660	0.6000	3.973	0.3626	0.0936	131	-0.86	003.14	0
003	Fat, Crude (%)	0027	3.664	0.0570	3.973	0.3626	0.0936	131	-0.85	003.09	0
003	Fat, Crude (%)	0265	3.665	0.0300	3.973	0.3626	0.0936	131	-0.85	003.14	0
003	Fat, Crude (%)	2155	3.670	0.0690	3.973	0.3626	0.0936	131	-0.84	003.10	0
003	Fat, Crude (%)	2073	3.670	0.0400	3.973	0.3626	0.0936	131	-0.84	003.10	0
003	Fat, Crude (%)	2089	3.685	0.1300	3.973	0.3626	0.0936	131	-0.79	003.13	0
003	Fat, Crude (%)	0171	3.690	0.7000	3.973	0.3626	0.0936	131	-0.78	003.14	0
003	Fat, Crude (%)	2295	3.695	0.0100	3.973	0.3626	0.0936	131	-0.77	003.12	0
003	Fat, Crude (%)	0062	3.700	0.1720	3.973	0.3626	0.0936	131	-0.75	003.10	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
003	Fat, Crude (%)	0098	3.700	0.0800	3.973	0.3626	0.0936	131	-0.75	003.10	0
003	Fat, Crude (%)	0357	3.700	0.2000	3.973	0.3626	0.0936	131	-0.75	003.14	0
003	Fat, Crude (%)	0175	3.705	0.4700	3.973	0.3626	0.0936	131	-0.74	003.14	0
003	Fat, Crude (%)	2246	3.720	0.0800	3.973	0.3626	0.0936	131	-0.70	003.13	0
003	Fat, Crude (%)	2062	3.724	0.0333	3.973	0.3626	0.0936	131	-0.69	003.12	0
003	Fat, Crude (%)	0693	3.732	0.0446	3.973	0.3626	0.0936	131	-0.67	003.10	0
003	Fat, Crude (%)	0407	3.733	0.0494	3.973	0.3626	0.0936	131	-0.66	003.14	0
003	Fat, Crude (%)	0202	3.735	0.2300	3.973	0.3626	0.0936	131	-0.66	003.14	0
003	Fat, Crude (%)	0511	3.735	0.1500	3.973	0.3626	0.0936	131	-0.66	003.14	0
003	Fat, Crude (%)	0049	3.750	0.2600	3.973	0.3626	0.0936	131	-0.62	003.14	0
003	Fat, Crude (%)	0089	3.770	0.0000	3.973	0.3626	0.0936	131	-0.56	003.14	0
003	Fat, Crude (%)	0653	3.780	0.0000	3.973	0.3626	0.0936	131	-0.53	003.14	0
003	Fat, Crude (%)	0878	3.800	0.0600	3.973	0.3626	0.0936	131	-0.48	003.14	0
003	Fat, Crude (%)	0948	3.800	0.0800	3.973	0.3626	0.0936	131	-0.48	003.14	0
003	Fat, Crude (%)	0956	3.800	0.0000	3.973	0.3626	0.0936	131	-0.48	003.14	0
003	Fat, Crude (%)	0512	3.811	0.0640	3.973	0.3626	0.0936	131	-0.45	003.09	0
003	Fat, Crude (%)	0868	3.815	0.0500	3.973	0.3626	0.0936	131	-0.44	003.14	0
003	Fat, Crude (%)	2199	3.825	0.1700	3.973	0.3626	0.0936	131	-0.41	003.12	0
003	Fat, Crude (%)	0674	3.830	0.0000	3.973	0.3626	0.0936	131	-0.40	003.09	0
003	Fat, Crude (%)	0100	3.830	0.0600	3.973	0.3626	0.0936	131	-0.40	003.10	0
003	Fat, Crude (%)	0033	3.850	0.0000	3.973	0.3626	0.0936	131	-0.34	003.13	0
003	Fat, Crude (%)	0001	3.854	0.0380	3.973	0.3626	0.0936	131	-0.33	003.14	0
003	Fat, Crude (%)	0619	3.855	0.0100	3.973	0.3626	0.0936	131	-0.33	003.10	0
003	Fat, Crude (%)	0598	3.855	0.0100	3.973	0.3626	0.0936	131	-0.33	003.14	0
003	Fat, Crude (%)	0098	3.870	0.0400	3.973	0.3626	0.0936	131	-0.28	003.13	0
003	Fat, Crude (%)	2022	3.875	0.0100	3.973	0.3626	0.0936	131	-0.27	003.10	0
003	Fat, Crude (%)	2181	3.880	0.0800	3.973	0.3626	0.0936	131	-0.26	003.06	0
003	Fat, Crude (%)	0042	3.890	0.2800	3.973	0.3626	0.0936	131	-0.23	003.10	0
003	Fat, Crude (%)	0529	3.905	0.1100	3.973	0.3626	0.0936	131	-0.19	003.14	0
003	Fat, Crude (%)	0425	3.910	0.0400	3.973	0.3626	0.0936	131	-0.17	003.06	0
003	Fat, Crude (%)	0563	3.921	0.0224	3.973	0.3626	0.0936	131	-0.14	003.06	0
003	Fat, Crude (%)	0563	3.921	0.0224	3.973	0.3626	0.0936	131	-0.14	003.14	0
003	Fat, Crude (%)	0003	3.930	0.3200	3.973	0.3626	0.0936	131	-0.12	003.14	0
003	Fat, Crude (%)	0009	3.930	0.0200	3.973	0.3626	0.0936	131	-0.12	003.14	0
003	Fat, Crude (%)	0066	3.930	0.0400	3.973	0.3626	0.0936	131	-0.12	003.14	0
003	Fat, Crude (%)	0968	3.945	0.0250	3.973	0.3626	0.0936	131	-0.08	003.14	0
003	Fat, Crude (%)	0541	3.960	0.0800	3.973	0.3626	0.0936	131	-0.04	003.10	0
003	Fat, Crude (%)	2193	3.960	0.0200	3.973	0.3626	0.0936	131	-0.04	003.10	0
003	Fat, Crude (%)	2009	3.967	0.0046	3.973	0.3626	0.0936	131	-0.02	003.10	0
003	Fat, Crude (%)	2196	3.970	0.0000	3.973	0.3626	0.0936	131	-0.01	003.10	0
003	Fat, Crude (%)	2012	3.975	0.0100	3.973	0.3626	0.0936	131	0.00	003.09	0
003	Fat, Crude (%)	0881	3.987	0.1170	3.973	0.3626	0.0936	131	0.04	003.10	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
003	Fat, Crude (%)	0123	4.000	0.0200	3.973	0.3626	0.0936	131	0.07	003.13	0
003	Fat, Crude (%)	0853	4.000	0.6000	3.973	0.3626	0.0936	131	0.07	003.14	0
003	Fat, Crude (%)	0297	4.005	0.0500	3.973	0.3626	0.0936	131	0.09	003.06	0
003	Fat, Crude (%)	0675	4.005	0.0500	3.973	0.3626	0.0936	131	0.09	003.09	0
003	Fat, Crude (%)	0345	4.030	0.0400	3.973	0.3626	0.0936	131	0.16	003.00	0
003	Fat, Crude (%)	0294	4.030	0.0800	3.973	0.3626	0.0936	131	0.16	003.06	0
003	Fat, Crude (%)	0644	4.047	0.0020	3.973	0.3626	0.0936	131	0.20	003.06	0
003	Fat, Crude (%)	0619	4.050	0.0800	3.973	0.3626	0.0936	131	0.21	003.14	0
003	Fat, Crude (%)	0226	4.100	0.1400	3.973	0.3626	0.0936	131	0.35	003.09	0
003	Fat, Crude (%)	0689	4.100	0.0000	3.973	0.3626	0.0936	131	0.35	003.10	0
003	Fat, Crude (%)	0692	4.100	0.0000	3.973	0.3626	0.0936	131	0.35	003.14	0
003	Fat, Crude (%)	0918	4.115	0.0900	3.973	0.3626	0.0936	131	0.39	003.06	0
003	Fat, Crude (%)	0948	4.125	0.0500	3.973	0.3626	0.0936	131	0.42	003.09	0
003	Fat, Crude (%)	2006	4.125	0.0500	3.973	0.3626	0.0936	131	0.42	003.13	0
003	Fat, Crude (%)	0870	4.126	0.2639	3.973	0.3626	0.0936	131	0.42	003.10	0
003	Fat, Crude (%)	0229	4.137	0.0100	3.973	0.3626	0.0936	131	0.45	003.14	0
003	Fat, Crude (%)	0870	4.140	0.0108	3.973	0.3626	0.0936	131	0.46	003.13	0
003	Fat, Crude (%)	0510	4.150	0.1000	3.973	0.3626	0.0936	131	0.49	003.09	0
003	Fat, Crude (%)	0278	4.150	0.1000	3.973	0.3626	0.0936	131	0.49	003.14	0
003	Fat, Crude (%)	0354	4.155	0.0100	3.973	0.3626	0.0936	131	0.50	003.09	0
003	Fat, Crude (%)	2033	4.160	0.0200	3.973	0.3626	0.0936	131	0.52	003.06	0
003	Fat, Crude (%)	2053	4.180	0.0200	3.973	0.3626	0.0936	131	0.57	003.00	0
003	Fat, Crude (%)	0148	4.185	0.3700	3.973	0.3626	0.0936	131	0.58	003.14	0
003	Fat, Crude (%)	0229	4.221	0.0330	3.973	0.3626	0.0936	131	0.68	003.06	0
003	Fat, Crude (%)	0098	4.235	0.3700	3.973	0.3626	0.0936	131	0.72	003.09	0
003	Fat, Crude (%)	2192	4.245	0.0900	3.973	0.3626	0.0936	131	0.75	003.00	0
003	Fat, Crude (%)	0685	4.245	0.0100	3.973	0.3626	0.0936	131	0.75	003.09	0
003	Fat, Crude (%)	0689	4.250	0.1000	3.973	0.3626	0.0936	131	0.76	003.06	0
003	Fat, Crude (%)	0366	4.250	0.1000	3.973	0.3626	0.0936	131	0.76	003.10	0
003	Fat, Crude (%)	2200	4.250	0.1000	3.973	0.3626	0.0936	131	0.76	003.12	0
003	Fat, Crude (%)	0164	4.260	0.0200	3.973	0.3626	0.0936	131	0.79	003.01	0
003	Fat, Crude (%)	0590	4.260	0.1800	3.973	0.3626	0.0936	131	0.79	003.09	0
003	Fat, Crude (%)	0004	4.260	0.0200	3.973	0.3626	0.0936	131	0.79	003.14	0
003	Fat, Crude (%)	0687	4.295	0.2100	3.973	0.3626	0.0936	131	0.89	003.06	0
003	Fat, Crude (%)	0413	4.300	0.2000	3.973	0.3626	0.0936	131	0.90	003.14	0
003	Fat, Crude (%)	0004	4.310	0.2000	3.973	0.3626	0.0936	131	0.93	003.09	0
003	Fat, Crude (%)	0309	4.310	0.0235	3.973	0.3626	0.0936	131	0.93	003.00	0
003	Fat, Crude (%)	0148	4.320	0.0600	3.973	0.3626	0.0936	131	0.96	003.10	0
003	Fat, Crude (%)	0169	4.335	0.0700	3.973	0.3626	0.0936	131	1.00	003.06	0
003	Fat, Crude (%)	2109	4.340	0.0000	3.973	0.3626	0.0936	131	1.01	003.00	0
003	Fat, Crude (%)	0910	4.365	0.2100	3.973	0.3626	0.0936	131	1.08	003.06	0
003	Fat, Crude (%)	0553	4.395	0.0300	3.973	0.3626	0.0936	131	1.16	003.11	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
003	Fat, Crude (%)	0726	4.400	0.0200	3.973	0.3626	0.0936	131	1.18	003.14	0
003	Fat, Crude (%)	0353	4.435	0.1700	3.973	0.3626	0.0936	131	1.27	003.09	0
003	Fat, Crude (%)	0175	4.445	0.1700	3.973	0.3626	0.0936	131	1.30	003.00	0
003	Fat, Crude (%)	0682	4.470	0.0000	3.973	0.3626	0.0936	131	1.37	003.06	0
003	Fat, Crude (%)	0870	4.477	0.0512	3.973	0.3626	0.0936	131	1.39	003.09	0
003	Fat, Crude (%)	0884	4.500	0.2000	3.973	0.3626	0.0936	131	1.45	003.06	0
003	Fat, Crude (%)	0405	4.510	0.0200	3.973	0.3626	0.0936	131	1.48	003.11	0
003	Fat, Crude (%)	0199	4.520	0.0400	3.973	0.3626	0.0936	131	1.51	003.06	0
003	Fat, Crude (%)	0190	4.630	0.0200	3.973	0.3626	0.0936	131	1.81	003.14	0
003	Fat, Crude (%)	0676	4.640	0.1600	3.973	0.3626	0.0936	131	1.84	003.10	0
003	Fat, Crude (%)	0032	4.700	0.1000	3.973	0.3626	0.0936	131	2.00	003.00	0
003	Fat, Crude (%)	0889	4.920	0.0400	3.973	0.3626	0.0936	131	2.61	003.11	0
003	Fat, Crude (%)	0297	5.125	0.0500	3.973	0.3626	0.0936	131	3.18	003.11	0
003	Fat, Crude (%)	0682	5.240	0.0000	3.973	0.3626	0.0936	131	3.49	003.14	0
003	Fat, Crude (%)	0852	5.500	0.0400	3.973	0.3626	0.0936	131	4.21	003.11	0
003	Fat, Crude (%)	0171	5.555	0.1500	3.973	0.3626	0.0936	131	4.36	003.12	0
003	Fat, Crude (%)	2129	7.313	0.0610	3.973	0.3626	0.0936	131	9.21	003.14	0
003	Fat, Crude (%)	0889	2.940	1.880	3.973	0.3626	0.0936	131	-2.85	003.10	1
003	Fat, Crude (%)	0074	4.535	0.7900	3.973	0.3626	0.0936	131	1.55	003.14	1
003	Fat, Crude (%)	2129	8.405	0.0730	3.973	0.3626	0.0936	131	12.22	003.11	2
004	Fiber, Crude (%)	2129	5.012	0.0810	13.15	0.8062	0.2583	110	-10.09	004.99	0
004	Fiber, Crude (%)	2129	5.581	0.3930	13.15	0.8062	0.2583	110	-9.39	004.11	0
004	Fiber, Crude (%)	2199	10.10	0.4000	13.15	0.8062	0.2583	110	-3.78	004.06	0
004	Fiber, Crude (%)	0675	10.53	0.4700	13.15	0.8062	0.2583	110	-3.25	004.07	0
004	Fiber, Crude (%)	0675	10.87	0.0900	13.15	0.8062	0.2583	110	-2.83	004.06	0
004	Fiber, Crude (%)	2146	11.26	0.0500	13.15	0.8062	0.2583	110	-2.35	004.99	0
004	Fiber, Crude (%)	0682	11.30	0.0000	13.15	0.8062	0.2583	110	-2.29	004.07	0
004	Fiber, Crude (%)	0028	11.70	0.0000	13.15	0.8062	0.2583	110	-1.79	004.07	0
004	Fiber, Crude (%)	0144	11.86	0.1600	13.15	0.8062	0.2583	110	-1.60	004.07	0
004	Fiber, Crude (%)	2246	12.11	0.1200	13.15	0.8062	0.2583	110	-1.29	004.03	0
004	Fiber, Crude (%)	2200	12.15	0.1000	13.15	0.8062	0.2583	110	-1.24	004.06	0
004	Fiber, Crude (%)	0366	12.20	1.000	13.15	0.8062	0.2583	110	-1.17	004.01	0
004	Fiber, Crude (%)	0354	12.25	0.0600	13.15	0.8062	0.2583	110	-1.11	004.00	0
004	Fiber, Crude (%)	0042	12.27	0.7800	13.15	0.8062	0.2583	110	-1.09	004.07	0
004	Fiber, Crude (%)	0910	12.30	0.0000	13.15	0.8062	0.2583	110	-1.05	004.07	0
004	Fiber, Crude (%)	0897	12.41	0.0600	13.15	0.8062	0.2583	110	-0.91	004.06	0
004	Fiber, Crude (%)	0004	12.44	0.2400	13.15	0.8062	0.2583	110	-0.88	004.07	0
004	Fiber, Crude (%)	0884	12.45	0.5000	13.15	0.8062	0.2583	110	-0.86	004.00	0
004	Fiber, Crude (%)	0644	12.45	0.4210	13.15	0.8062	0.2583	110	-0.86	004.07	0
004	Fiber, Crude (%)	0848	12.48	0.0000	13.15	0.8062	0.2583	110	-0.83	004.07	0
004	Fiber, Crude (%)	2073	12.50	0.0700	13.15	0.8062	0.2583	110	-0.81	004.00	0
004	Fiber, Crude (%)	2009	12.50	0.0130	13.15	0.8062	0.2583	110	-0.80	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber, Crude (%)	0278	12.50	0.2000	13.15	0.8062	0.2583	110	-0.80	004.07	0
004	Fiber, Crude (%)	0510	12.50	0.0000	13.15	0.8062	0.2583	110	-0.80	004.07	0
004	Fiber, Crude (%)	2053	12.51	0.0800	13.15	0.8062	0.2583	110	-0.79	004.07	0
004	Fiber, Crude (%)	0202	12.56	0.4000	13.15	0.8062	0.2583	110	-0.73	004.07	0
004	Fiber, Crude (%)	0413	12.60	0.8000	13.15	0.8062	0.2583	110	-0.68	004.07	0
004	Fiber, Crude (%)	0956	12.60	0.4000	13.15	0.8062	0.2583	110	-0.68	004.07	0
004	Fiber, Crude (%)	2109	12.62	0.1800	13.15	0.8062	0.2583	110	-0.65	004.07	0
004	Fiber, Crude (%)	0294	12.65	0.1000	13.15	0.8062	0.2583	110	-0.62	004.07	0
004	Fiber, Crude (%)	0098	12.66	0.5500	13.15	0.8062	0.2583	110	-0.61	004.06	0
004	Fiber, Crude (%)	2259	12.71	0.2440	13.15	0.8062	0.2583	110	-0.54	004.07	0
004	Fiber, Crude (%)	0019	12.76	0.1900	13.15	0.8062	0.2583	110	-0.49	004.07	0
004	Fiber, Crude (%)	0353	12.77	0.0300	13.15	0.8062	0.2583	110	-0.47	004.03	0
004	Fiber, Crude (%)	0098	12.80	0.2300	13.15	0.8062	0.2583	110	-0.44	004.07	0
004	Fiber, Crude (%)	0263	12.80	0.1020	13.15	0.8062	0.2583	110	-0.43	004.07	0
004	Fiber, Crude (%)	0017	12.80	0.2000	13.15	0.8062	0.2583	110	-0.43	004.00	0
004	Fiber, Crude (%)	2304	12.81	0.3000	13.15	0.8062	0.2583	110	-0.42	004.07	0
004	Fiber, Crude (%)	2152	12.83	0.0600	13.15	0.8062	0.2583	110	-0.39	004.06	0
004	Fiber, Crude (%)	0553	12.89	0.1800	13.15	0.8062	0.2583	110	-0.32	004.11	0
004	Fiber, Crude (%)	0692	12.90	0.0000	13.15	0.8062	0.2583	110	-0.31	004.07	0
004	Fiber, Crude (%)	0674	12.95	0.0000	13.15	0.8062	0.2583	110	-0.24	004.06	0
004	Fiber, Crude (%)	0190	12.95	0.1000	13.15	0.8062	0.2583	110	-0.24	004.07	0
004	Fiber, Crude (%)	0878	12.96	0.0900	13.15	0.8062	0.2583	110	-0.24	004.06	0
004	Fiber, Crude (%)	0708	12.96	0.0500	13.15	0.8062	0.2583	110	-0.24	004.07	0
004	Fiber, Crude (%)	2022	12.96	0.0000	13.15	0.8062	0.2583	110	-0.23	004.06	0
004	Fiber, Crude (%)	0512	12.97	0.0700	13.15	0.8062	0.2583	110	-0.22	004.06	0
004	Fiber, Crude (%)	0689	13.00	0.6000	13.15	0.8062	0.2583	110	-0.18	004.06	0
004	Fiber, Crude (%)	0876	13.00	0.0000	13.15	0.8062	0.2583	110	-0.18	004.07	0
004	Fiber, Crude (%)	2012	13.02	0.2800	13.15	0.8062	0.2583	110	-0.16	004.07	0
004	Fiber, Crude (%)	0948	13.07	0.3330	13.15	0.8062	0.2583	110	-0.10	004.06	0
004	Fiber, Crude (%)	0852	13.08	0.0300	13.15	0.8062	0.2583	110	-0.09	004.11	0
004	Fiber, Crude (%)	0425	13.10	0.0000	13.15	0.8062	0.2583	110	-0.06	004.00	0
004	Fiber, Crude (%)	0968	13.13	0.3930	13.15	0.8062	0.2583	110	-0.02	004.06	0
004	Fiber, Crude (%)	0868	13.14	0.0500	13.15	0.8062	0.2583	110	-0.01	004.06	0
004	Fiber, Crude (%)	0027	13.15	0.3510	13.15	0.8062	0.2583	110	0.00	004.06	0
004	Fiber, Crude (%)	0889	13.15	0.2050	13.15	0.8062	0.2583	110	0.01	004.11	0
004	Fiber, Crude (%)	0870	13.18	0.2084	13.15	0.8062	0.2583	110	0.04	004.07	0
004	Fiber, Crude (%)	0968	13.19	0.4880	13.15	0.8062	0.2583	110	0.05	004.07	0
004	Fiber, Crude (%)	0405	13.20	0.0700	13.15	0.8062	0.2583	110	0.06	004.11	0
004	Fiber, Crude (%)	0563	13.20	0.1079	13.15	0.8062	0.2583	110	0.07	004.07	0
004	Fiber, Crude (%)	0035	13.23	0.2300	13.15	0.8062	0.2583	110	0.10	004.07	0
004	Fiber, Crude (%)	0164	13.25	0.3000	13.15	0.8062	0.2583	110	0.13	004.00	0
004	Fiber, Crude (%)	0169	13.25	0.1000	13.15	0.8062	0.2583	110	0.13	004.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber, Crude (%)	0693	13.26	0.7858	13.15	0.8062	0.2583	110	0.14	004.07	0
004	Fiber, Crude (%)	0009	13.30	0.3800	13.15	0.8062	0.2583	110	0.19	004.07	0
004	Fiber, Crude (%)	0265	13.30	0.6000	13.15	0.8062	0.2583	110	0.19	004.07	0
004	Fiber, Crude (%)	0175	13.32	0.6200	13.15	0.8062	0.2583	110	0.22	004.00	0
004	Fiber, Crude (%)	0881	13.33	0.0180	13.15	0.8062	0.2583	110	0.23	004.06	0
004	Fiber, Crude (%)	0083	13.34	0.1700	13.15	0.8062	0.2583	110	0.23	004.07	0
004	Fiber, Crude (%)	0100	13.35	1.250	13.15	0.8062	0.2583	110	0.25	004.07	0
004	Fiber, Crude (%)	2196	13.36	0.0000	13.15	0.8062	0.2583	110	0.27	004.07	0
004	Fiber, Crude (%)	0598	13.37	0.0500	13.15	0.8062	0.2583	110	0.27	004.07	0
004	Fiber, Crude (%)	0726	13.38	0.0000	13.15	0.8062	0.2583	110	0.29	004.07	0
004	Fiber, Crude (%)	0354	13.42	0.0400	13.15	0.8062	0.2583	110	0.34	004.07	0
004	Fiber, Crude (%)	0948	13.42	0.1720	13.15	0.8062	0.2583	110	0.34	004.07	0
004	Fiber, Crude (%)	2062	13.45	0.1751	13.15	0.8062	0.2583	110	0.37	004.06	0
004	Fiber, Crude (%)	0345	13.45	0.1000	13.15	0.8062	0.2583	110	0.38	004.00	0
004	Fiber, Crude (%)	0034	13.46	0.1900	13.15	0.8062	0.2583	110	0.38	004.07	0
004	Fiber, Crude (%)	0309	13.54	0.2452	13.15	0.8062	0.2583	110	0.49	004.00	0
004	Fiber, Crude (%)	2181	13.55	0.3000	13.15	0.8062	0.2583	110	0.50	004.07	0
004	Fiber, Crude (%)	0242	13.59	0.7300	13.15	0.8062	0.2583	110	0.54	004.07	0
004	Fiber, Crude (%)	0123	13.61	0.2200	13.15	0.8062	0.2583	110	0.58	004.07	0
004	Fiber, Crude (%)	0619	13.65	0.1000	13.15	0.8062	0.2583	110	0.62	004.07	0
004	Fiber, Crude (%)	0199	13.66	0.0600	13.15	0.8062	0.2583	110	0.64	004.06	0
004	Fiber, Crude (%)	2193	13.67	0.0400	13.15	0.8062	0.2583	110	0.65	004.07	0
004	Fiber, Crude (%)	0529	13.67	0.0800	13.15	0.8062	0.2583	110	0.65	004.07	0
004	Fiber, Crude (%)	0407	13.70	0.5761	13.15	0.8062	0.2583	110	0.69	004.07	0
004	Fiber, Crude (%)	0941	13.74	0.3800	13.15	0.8062	0.2583	110	0.74	004.06	0
004	Fiber, Crude (%)	0089	13.79	0.1000	13.15	0.8062	0.2583	110	0.80	004.07	0
004	Fiber, Crude (%)	0066	13.85	0.3400	13.15	0.8062	0.2583	110	0.87	004.07	0
004	Fiber, Crude (%)	0074	13.97	0.3700	13.15	0.8062	0.2583	110	1.02	004.07	0
004	Fiber, Crude (%)	2076	14.03	0.4959	13.15	0.8062	0.2583	110	1.10	004.00	0
004	Fiber, Crude (%)	0918	14.07	0.0900	13.15	0.8062	0.2583	110	1.14	004.06	0
004	Fiber, Crude (%)	0643	14.21	0.5000	13.15	0.8062	0.2583	110	1.32	004.07	0
004	Fiber, Crude (%)	0505	14.22	0.4600	13.15	0.8062	0.2583	110	1.33	004.07	0
004	Fiber, Crude (%)	0581	14.25	0.1000	13.15	0.8062	0.2583	110	1.37	004.07	0
004	Fiber, Crude (%)	0226	14.34	0.5400	13.15	0.8062	0.2583	110	1.48	004.00	0
004	Fiber, Crude (%)	0229	14.39	0.1300	13.15	0.8062	0.2583	110	1.54	004.07	0
004	Fiber, Crude (%)	0298	14.40	0.2000	13.15	0.8062	0.2583	110	1.56	004.07	0
004	Fiber, Crude (%)	2192	14.43	0.0000	13.15	0.8062	0.2583	110	1.59	004.03	0
004	Fiber, Crude (%)	0032	14.44	0.9700	13.15	0.8062	0.2583	110	1.60	004.07	0
004	Fiber, Crude (%)	2006	14.49	0.6000	13.15	0.8062	0.2583	110	1.67	004.06	0
004	Fiber, Crude (%)	2033	14.50	0.2000	13.15	0.8062	0.2583	110	1.68	004.07	0
004	Fiber, Crude (%)	0536	14.63	0.9300	13.15	0.8062	0.2583	110	1.83	004.07	0
004	Fiber, Crude (%)	0511	14.65	0.5000	13.15	0.8062	0.2583	110	1.87	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
004	Fiber, Crude (%)	0893	14.87	0.1400	13.15	0.8062	0.2583	110	2.14	004.06	0
004	Fiber, Crude (%)	0096	15.29	0.2000	13.15	0.8062	0.2583	110	2.66	004.07	0
004	Fiber, Crude (%)	0676	15.36	0.5100	13.15	0.8062	0.2583	110	2.74	004.06	0
004	Fiber, Crude (%)	2089	21.45	0.4500	13.15	0.8062	0.2583	110	10.29	004.03	0
004	Fiber, Crude (%)	0520	14.06	1.330	13.15	0.8062	0.2583	110	1.13	004.07	1
004	Fiber, Crude (%)	0171	14.42	1.520	13.15	0.8062	0.2583	110	1.58	004.00	1
005	Ash (%)	0297	7.910	0.0600	13.72	0.2871	0.1438	133	-20.24	005.11	0
005	Ash (%)	0889	9.820	0.2700	13.72	0.2871	0.1438	133	-13.59	005.11	0
005	Ash (%)	2081	12.01	0.9000	13.72	0.2871	0.1438	133	-5.96	005.99	0
005	Ash (%)	0229	12.77	0.2900	13.72	0.2871	0.1438	133	-3.33	005.00	0
005	Ash (%)	0674	12.90	0.0000	13.72	0.2871	0.1438	133	-2.86	005.00	0
005	Ash (%)	0425	13.01	0.0900	13.72	0.2871	0.1438	133	-2.50	005.00	0
005	Ash (%)	0038	13.10	0.5500	13.72	0.2871	0.1438	133	-2.18	005.05	0
005	Ash (%)	2033	13.10	0.0000	13.72	0.2871	0.1438	133	-2.17	005.00	0
005	Ash (%)	0512	13.20	0.7000	13.72	0.2871	0.1438	133	-1.82	005.00	0
005	Ash (%)	0169	13.23	0.0600	13.72	0.2871	0.1438	133	-1.71	005.00	0
005	Ash (%)	0620	13.29	0.4339	13.72	0.2871	0.1438	133	-1.49	005.00	0
005	Ash (%)	0357	13.30	0.2000	13.72	0.2871	0.1438	133	-1.47	005.00	0
005	Ash (%)	2199	13.35	0.1000	13.72	0.2871	0.1438	133	-1.30	005.00	0
005	Ash (%)	0953	13.39	0.0100	13.72	0.2871	0.1438	133	-1.17	005.00	0
005	Ash (%)	0164	13.40	0.0100	13.72	0.2871	0.1438	133	-1.14	005.00	0
005	Ash (%)	0956	13.40	0.0000	13.72	0.2871	0.1438	133	-1.12	005.00	0
005	Ash (%)	0510	13.43	0.1700	13.72	0.2871	0.1438	133	-1.03	005.00	0
005	Ash (%)	0083	13.43	0.0400	13.72	0.2871	0.1438	133	-1.02	005.00	0
005	Ash (%)	2302	13.45	0.2500	13.72	0.2871	0.1438	133	-0.96	005.00	0
005	Ash (%)	0574	13.45	0.1000	13.72	0.2871	0.1438	133	-0.95	005.00	0
005	Ash (%)	0144	13.47	0.1900	13.72	0.2871	0.1438	133	-0.89	005.00	0
005	Ash (%)	2006	13.47	0.0300	13.72	0.2871	0.1438	133	-0.89	005.00	0
005	Ash (%)	0175	13.48	0.0900	13.72	0.2871	0.1438	133	-0.86	005.00	0
005	Ash (%)	0511	13.48	0.0700	13.72	0.2871	0.1438	133	-0.86	005.00	0
005	Ash (%)	0539	13.48	0.1600	13.72	0.2871	0.1438	133	-0.84	005.00	0
005	Ash (%)	0366	13.50	0.0000	13.72	0.2871	0.1438	133	-0.77	005.00	0
005	Ash (%)	0003	13.50	0.0000	13.72	0.2871	0.1438	133	-0.77	005.05	0
005	Ash (%)	0027	13.50	0.0320	13.72	0.2871	0.1438	133	-0.77	005.00	0
005	Ash (%)	0726	13.52	0.0300	13.72	0.2871	0.1438	133	-0.72	005.05	0
005	Ash (%)	0354	13.53	0.0300	13.72	0.2871	0.1438	133	-0.69	005.00	0
005	Ash (%)	0358	13.53	0.0000	13.72	0.2871	0.1438	133	-0.67	005.00	0
005	Ash (%)	0035	13.54	0.1100	13.72	0.2871	0.1438	133	-0.65	005.00	0
005	Ash (%)	0689	13.54	0.0400	13.72	0.2871	0.1438	133	-0.63	005.05	0
005	Ash (%)	0066	13.55	0.0700	13.72	0.2871	0.1438	133	-0.62	005.05	0
005	Ash (%)	0870	13.55	0.2185	13.72	0.2871	0.1438	133	-0.60	005.00	0
005	Ash (%)	0004	13.55	0.0800	13.72	0.2871	0.1438	133	-0.60	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
005	Ash (%)	0019	13.55	0.1000	13.72	0.2871	0.1438	133	-0.60	005.00	0
005	Ash (%)	0148	13.55	0.5000	13.72	0.2871	0.1438	133	-0.60	005.00	0
005	Ash (%)	0171	13.57	0.0700	13.72	0.2871	0.1438	133	-0.55	005.00	0
005	Ash (%)	2304	13.57	0.0500	13.72	0.2871	0.1438	133	-0.55	005.03	0
005	Ash (%)	0675	13.58	0.1400	13.72	0.2871	0.1438	133	-0.49	005.00	0
005	Ash (%)	0049	13.59	0.3200	13.72	0.2871	0.1438	133	-0.46	005.00	0
005	Ash (%)	0278	13.59	0.0600	13.72	0.2871	0.1438	133	-0.46	005.00	0
005	Ash (%)	0001	13.59	0.1057	13.72	0.2871	0.1438	133	-0.45	005.00	0
005	Ash (%)	0353	13.60	0.1900	13.72	0.2871	0.1438	133	-0.44	005.00	0
005	Ash (%)	2200	13.60	0.4000	13.72	0.2871	0.1438	133	-0.42	005.00	0
005	Ash (%)	0893	13.62	0.0500	13.72	0.2871	0.1438	133	-0.37	005.05	0
005	Ash (%)	0098	13.62	0.0000	13.72	0.2871	0.1438	133	-0.35	005.00	0
005	Ash (%)	0622	13.63	0.0012	13.72	0.2871	0.1438	133	-0.32	005.00	0
005	Ash (%)	0297	13.63	0.7400	13.72	0.2871	0.1438	133	-0.32	005.00	0
005	Ash (%)	0298	13.64	0.0300	13.72	0.2871	0.1438	133	-0.30	005.00	0
005	Ash (%)	0876	13.65	0.1000	13.72	0.2871	0.1438	133	-0.25	005.00	0
005	Ash (%)	0912	13.65	0.1000	13.72	0.2871	0.1438	133	-0.25	005.03	0
005	Ash (%)	0589	13.65	0.0400	13.72	0.2871	0.1438	133	-0.25	005.00	0
005	Ash (%)	0918	13.65	0.0000	13.72	0.2871	0.1438	133	-0.25	005.00	0
005	Ash (%)	0541	13.66	0.3200	13.72	0.2871	0.1438	133	-0.22	005.00	0
005	Ash (%)	2022	13.67	0.0500	13.72	0.2871	0.1438	133	-0.20	005.00	0
005	Ash (%)	0968	13.68	0.0080	13.72	0.2871	0.1438	133	-0.16	005.00	0
005	Ash (%)	2181	13.69	0.0100	13.72	0.2871	0.1438	133	-0.13	005.00	0
005	Ash (%)	2109	13.69	0.0200	13.72	0.2871	0.1438	133	-0.11	005.05	0
005	Ash (%)	0618	13.69	0.0600	13.72	0.2871	0.1438	133	-0.11	005.00	0
005	Ash (%)	2073	13.69	0.0800	13.72	0.2871	0.1438	133	-0.11	005.00	0
005	Ash (%)	0263	13.69	0.0202	13.72	0.2871	0.1438	133	-0.10	005.05	0
005	Ash (%)	0042	13.70	0.0300	13.72	0.2871	0.1438	133	-0.09	005.00	0
005	Ash (%)	0265	13.70	0.2000	13.72	0.2871	0.1438	133	-0.08	005.00	0
005	Ash (%)	0853	13.70	0.2000	13.72	0.2871	0.1438	133	-0.08	005.00	0
005	Ash (%)	0693	13.70	0.0185	13.72	0.2871	0.1438	133	-0.07	005.00	0
005	Ash (%)	0682	13.71	0.0000	13.72	0.2871	0.1438	133	-0.04	005.00	0
005	Ash (%)	0968	13.71	0.0260	13.72	0.2871	0.1438	133	-0.03	005.02	0
005	Ash (%)	0062	13.72	0.1510	13.72	0.2871	0.1438	133	0.01	005.00	0
005	Ash (%)	0598	13.73	0.1300	13.72	0.2871	0.1438	133	0.01	005.00	0
005	Ash (%)	0644	13.73	0.0800	13.72	0.2871	0.1438	133	0.03	005.02	0
005	Ash (%)	0881	13.73	0.1570	13.72	0.2871	0.1438	133	0.04	005.00	0
005	Ash (%)	0939	13.74	0.2600	13.72	0.2871	0.1438	133	0.06	005.05	0
005	Ash (%)	2259	13.75	0.3150	13.72	0.2871	0.1438	133	0.09	005.00	0
005	Ash (%)	0536	13.76	0.4300	13.72	0.2871	0.1438	133	0.12	005.99	0
005	Ash (%)	0646	13.77	0.0500	13.72	0.2871	0.1438	133	0.15	005.00	0
005	Ash (%)	2012	13.77	0.1300	13.72	0.2871	0.1438	133	0.15	005.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
005	Ash (%)	2196	13.78	0.0000	13.72	0.2871	0.1438	133	0.20	005.00	0
005	Ash (%)	0100	13.78	0.3000	13.72	0.2871	0.1438	133	0.20	005.00	0
005	Ash (%)	0848	13.78	0.2000	13.72	0.2871	0.1438	133	0.20	005.00	0
005	Ash (%)	0878	13.78	0.0600	13.72	0.2871	0.1438	133	0.20	005.05	0
005	Ash (%)	2295	13.79	0.0100	13.72	0.2871	0.1438	133	0.22	005.05	0
005	Ash (%)	2246	13.80	0.0900	13.72	0.2871	0.1438	133	0.25	005.05	0
005	Ash (%)	0563	13.80	0.0525	13.72	0.2871	0.1438	133	0.26	005.00	0
005	Ash (%)	0242	13.80	0.0200	13.72	0.2871	0.1438	133	0.27	005.00	0
005	Ash (%)	0413	13.80	0.2000	13.72	0.2871	0.1438	133	0.27	005.05	0
005	Ash (%)	0619	13.80	0.0000	13.72	0.2871	0.1438	133	0.27	005.05	0
005	Ash (%)	0202	13.81	0.2700	13.72	0.2871	0.1438	133	0.29	005.99	0
005	Ash (%)	0643	13.81	0.3400	13.72	0.2871	0.1438	133	0.31	005.00	0
005	Ash (%)	2193	13.82	0.0200	13.72	0.2871	0.1438	133	0.34	005.00	0
005	Ash (%)	0142	13.84	0.3400	13.72	0.2871	0.1438	133	0.41	005.00	0
005	Ash (%)	2053	13.85	0.1300	13.72	0.2871	0.1438	133	0.43	005.00	0
005	Ash (%)	0590	13.85	0.0500	13.72	0.2871	0.1438	133	0.43	005.05	0
005	Ash (%)	0294	13.85	0.1000	13.72	0.2871	0.1438	133	0.45	005.05	0
005	Ash (%)	0407	13.85	0.2427	13.72	0.2871	0.1438	133	0.46	005.00	0
005	Ash (%)	0653	13.86	0.0100	13.72	0.2871	0.1438	133	0.46	005.05	0
005	Ash (%)	2089	13.86	0.0000	13.72	0.2871	0.1438	133	0.48	005.00	0
005	Ash (%)	0948	13.87	0.0600	13.72	0.2871	0.1438	133	0.52	005.00	0
005	Ash (%)	2076	13.90	0.0492	13.72	0.2871	0.1438	133	0.61	005.00	0
005	Ash (%)	0596	13.90	0.0200	13.72	0.2871	0.1438	133	0.62	005.05	0
005	Ash (%)	0884	13.90	0.2000	13.72	0.2871	0.1438	133	0.62	005.99	0
005	Ash (%)	0914	13.91	0.2000	13.72	0.2871	0.1438	133	0.66	005.00	0
005	Ash (%)	0581	13.93	0.2100	13.72	0.2871	0.1438	133	0.71	005.00	0
005	Ash (%)	2144	13.93	0.0400	13.72	0.2871	0.1438	133	0.72	005.05	0
005	Ash (%)	0017	13.95	0.1000	13.72	0.2871	0.1438	133	0.79	005.00	0
005	Ash (%)	0868	13.97	0.0000	13.72	0.2871	0.1438	133	0.86	005.05	0
005	Ash (%)	0139	13.99	0.0700	13.72	0.2871	0.1438	133	0.92	005.00	0
005	Ash (%)	0505	13.99	0.0500	13.72	0.2871	0.1438	133	0.92	005.00	0
005	Ash (%)	0692	14.00	0.0000	13.72	0.2871	0.1438	133	0.97	005.05	0
005	Ash (%)	0852	14.01	0.0500	13.72	0.2871	0.1438	133	0.99	005.00	0
005	Ash (%)	2146	14.02	0.0500	13.72	0.2871	0.1438	133	1.02	005.05	0
005	Ash (%)	2150	14.02	0.1600	13.72	0.2871	0.1438	133	1.04	005.00	0
005	Ash (%)	0897	14.04	0.1800	13.72	0.2871	0.1438	133	1.11	005.05	0
005	Ash (%)	0309	14.05	0.0600	13.72	0.2871	0.1438	133	1.14	005.00	0
005	Ash (%)	2009	14.06	0.0070	13.72	0.2871	0.1438	133	1.17	005.05	0
005	Ash (%)	0199	14.10	0.0500	13.72	0.2871	0.1438	133	1.30	005.00	0
005	Ash (%)	0685	14.15	0.0100	13.72	0.2871	0.1438	133	1.47	005.05	0
005	Ash (%)	0610	14.15	0.1000	13.72	0.2871	0.1438	133	1.49	005.05	0
005	Ash (%)	0652	14.15	0.1000	13.72	0.2871	0.1438	133	1.49	005.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
005	Ash (%)	0921	14.17	0.1200	13.72	0.2871	0.1438	133	1.56	005.05	0
005	Ash (%)	0910	14.20	0.2000	13.72	0.2871	0.1438	133	1.67	005.99	0
005	Ash (%)	2062	14.23	0.0092	13.72	0.2871	0.1438	133	1.78	005.05	0
005	Ash (%)	0034	14.24	0.4342	13.72	0.2871	0.1438	133	1.80	005.05	0
005	Ash (%)	0047	14.30	0.2000	13.72	0.2871	0.1438	133	2.01	005.00	0
005	Ash (%)	0520	14.32	0.0400	13.72	0.2871	0.1438	133	2.08	005.00	0
005	Ash (%)	0529	14.45	1.600	13.72	0.2871	0.1438	133	2.54	005.00	0
005	Ash (%)	0123	14.50	0.4000	13.72	0.2871	0.1438	133	2.71	005.00	0
005	Ash (%)	0676	14.59	0.2100	13.72	0.2871	0.1438	133	3.01	005.99	0
005	Ash (%)	0345	14.73	0.1500	13.72	0.2871	0.1438	133	3.49	005.00	0
005	Ash (%)	2192	14.82	0.1400	13.72	0.2871	0.1438	133	3.82	005.05	0
005	Ash (%)	0190	14.90	0.1600	13.72	0.2871	0.1438	133	4.10	005.05	0
005	Ash (%)	0226	15.22	0.1400	13.72	0.2871	0.1438	133	5.22	005.00	0
005	Ash (%)	0889	10.24	6.710	13.72	0.2871	0.1438	133	-12.14	005.05	1
005	Ash (%)	0852	6.445	0.0100	13.72	0.2871	0.1438	133	-25.35	005.11	2
005	Ash (%)	2129	6.720	0.0660	13.72	0.2871	0.1438	133	-24.39	005.05	2
006	Total Sugars (%)	0921	4.390	0.1600	5.045	0.4846	0.1293	4		006.00	0
006	Total Sugars (%)	0910	4.990	0.1000	5.045	0.4846	0.1293	4		006.00	0
006	Total Sugars (%)	0407	5.301	0.0570	5.045	0.4846	0.1293	4		006.00	0
006	Total Sugars (%)	0956	5.500	0.2000	5.045	0.4846	0.1293	4		006.99	0
008	Fiber, Acid Detergent (%)	2129	5.790	0.3260	18.07	1.955	0.3220	57	-6.28	008.08	0
008	Fiber, Acid Detergent (%)	0297	8.970	1.060	18.07	1.955	0.3220	57	-4.65	008.08	0
008	Fiber, Acid Detergent (%)	0956	14.70	1.200	18.07	1.955	0.3220	57	-1.72	008.08	0
008	Fiber, Acid Detergent (%)	0968	14.84	0.1560	18.07	1.955	0.3220	57	-1.65	008.08	0
008	Fiber, Acid Detergent (%)	0354	14.89	0.0200	18.07	1.955	0.3220	57	-1.63	008.08	0
008	Fiber, Acid Detergent (%)	0590	14.90	0.4000	18.07	1.955	0.3220	57	-1.62	008.08	0
008	Fiber, Acid Detergent (%)	0027	15.56	0.3560	18.07	1.955	0.3220	57	-1.28	008.99	0
008	Fiber, Acid Detergent (%)	0676	15.64	0.0200	18.07	1.955	0.3220	57	-1.24	008.02	0
008	Fiber, Acid Detergent (%)	0848	15.87	0.0100	18.07	1.955	0.3220	57	-1.13	008.08	0
008	Fiber, Acid Detergent (%)	0948	15.89	0.3600	18.07	1.955	0.3220	57	-1.12	008.02	0
008	Fiber, Acid Detergent (%)	0675	16.07	0.0700	18.07	1.955	0.3220	57	-1.03	008.08	0
008	Fiber, Acid Detergent (%)	0941	16.07	0.3900	18.07	1.955	0.3220	57	-1.03	008.99	0
008	Fiber, Acid Detergent (%)	0190	16.27	0.5900	18.07	1.955	0.3220	57	-0.92	008.08	0
008	Fiber, Acid Detergent (%)	0948	16.29	0.4960	18.07	1.955	0.3220	57	-0.91	008.08	0
008	Fiber, Acid Detergent (%)	0675	16.45	0.1900	18.07	1.955	0.3220	57	-0.83	008.02	0
008	Fiber, Acid Detergent (%)	0596	16.67	0.0900	18.07	1.955	0.3220	57	-0.72	008.08	0
008	Fiber, Acid Detergent (%)	0049	16.99	0.8700	18.07	1.955	0.3220	57	-0.56	008.08	0
008	Fiber, Acid Detergent (%)	0689	17.20	0.0000	18.07	1.955	0.3220	57	-0.45	008.02	0
008	Fiber, Acid Detergent (%)	2012	17.27	0.7500	18.07	1.955	0.3220	57	-0.41	008.08	0
008	Fiber, Acid Detergent (%)	0878	17.29	0.3900	18.07	1.955	0.3220	57	-0.40	008.08	0
008	Fiber, Acid Detergent (%)	0870	17.64	0.3559	18.07	1.955	0.3220	57	-0.22	008.08	0
008	Fiber, Acid Detergent (%)	0042	17.92	0.1600	18.07	1.955	0.3220	57	-0.08	008.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
008	Fiber, Acid Detergent (%)	0358	18.04	0.3900	18.07	1.955	0.3220	57	-0.02	008.08	0
008	Fiber, Acid Detergent (%)	2196	18.08	0.0000	18.07	1.955	0.3220	57	0.00	008.08	0
008	Fiber, Acid Detergent (%)	0098	18.21	0.4200	18.07	1.955	0.3220	57	0.07	008.02	0
008	Fiber, Acid Detergent (%)	0511	18.30	0.6000	18.07	1.955	0.3220	57	0.12	008.99	0
008	Fiber, Acid Detergent (%)	0407	18.38	0.2434	18.07	1.955	0.3220	57	0.16	008.08	0
008	Fiber, Acid Detergent (%)	0918	18.40	0.4393	18.07	1.955	0.3220	57	0.17	008.08	0
008	Fiber, Acid Detergent (%)	0536	18.47	0.7600	18.07	1.955	0.3220	57	0.20	008.08	0
008	Fiber, Acid Detergent (%)	0309	18.58	0.0821	18.07	1.955	0.3220	57	0.26	008.02	0
008	Fiber, Acid Detergent (%)	0510	18.75	0.1000	18.07	1.955	0.3220	57	0.35	008.08	0
008	Fiber, Acid Detergent (%)	0353	18.79	0.1000	18.07	1.955	0.3220	57	0.37	008.02	0
008	Fiber, Acid Detergent (%)	0581	18.80	0.2600	18.07	1.955	0.3220	57	0.37	008.08	0
008	Fiber, Acid Detergent (%)	0413	18.85	0.3000	18.07	1.955	0.3220	57	0.40	008.08	0
008	Fiber, Acid Detergent (%)	0912	18.90	0.4000	18.07	1.955	0.3220	57	0.42	008.08	0
008	Fiber, Acid Detergent (%)	0345	19.00	0.4000	18.07	1.955	0.3220	57	0.48	008.02	0
008	Fiber, Acid Detergent (%)	0004	19.02	0.0300	18.07	1.955	0.3220	57	0.48	008.08	0
008	Fiber, Acid Detergent (%)	0083	19.02	0.0800	18.07	1.955	0.3220	57	0.49	008.08	0
008	Fiber, Acid Detergent (%)	0265	19.25	0.1000	18.07	1.955	0.3220	57	0.60	008.05	0
008	Fiber, Acid Detergent (%)	0066	19.26	0.3100	18.07	1.955	0.3220	57	0.61	008.08	0
008	Fiber, Acid Detergent (%)	0035	19.29	0.4600	18.07	1.955	0.3220	57	0.62	008.08	0
008	Fiber, Acid Detergent (%)	0164	19.35	0.1000	18.07	1.955	0.3220	57	0.65	008.08	0
008	Fiber, Acid Detergent (%)	0294	19.35	0.1000	18.07	1.955	0.3220	57	0.65	008.08	0
008	Fiber, Acid Detergent (%)	0148	19.40	0.6000	18.07	1.955	0.3220	57	0.68	008.08	0
008	Fiber, Acid Detergent (%)	0001	19.41	0.0100	18.07	1.955	0.3220	57	0.68	008.08	0
008	Fiber, Acid Detergent (%)	0263	19.41	0.1296	18.07	1.955	0.3220	57	0.69	008.08	0
008	Fiber, Acid Detergent (%)	0619	19.45	0.1000	18.07	1.955	0.3220	57	0.71	008.08	0
008	Fiber, Acid Detergent (%)	2193	19.51	0.7400	18.07	1.955	0.3220	57	0.74	008.08	0
008	Fiber, Acid Detergent (%)	0037	19.61	0.0900	18.07	1.955	0.3220	57	0.78	008.08	0
008	Fiber, Acid Detergent (%)	0726	19.90	0.1200	18.07	1.955	0.3220	57	0.94	008.08	0
008	Fiber, Acid Detergent (%)	0226	19.92	0.4400	18.07	1.955	0.3220	57	0.95	008.02	0
008	Fiber, Acid Detergent (%)	0148	20.00	0.4000	18.07	1.955	0.3220	57	0.99	008.02	0
008	Fiber, Acid Detergent (%)	0278	20.40	0.0000	18.07	1.955	0.3220	57	1.19	008.08	0
008	Fiber, Acid Detergent (%)	2192	20.49	0.7700	18.07	1.955	0.3220	57	1.23	008.02	0
008	Fiber, Acid Detergent (%)	0357	20.75	0.3000	18.07	1.955	0.3220	57	1.37	008.08	0
008	Fiber, Acid Detergent (%)	0693	20.83	0.1314	18.07	1.955	0.3220	57	1.41	008.08	0
008	Fiber, Acid Detergent (%)	0646	21.83	0.5900	18.07	1.955	0.3220	57	1.92	008.08	0
008	Fiber, Acid Detergent (%)	0889	13.80	4.130	18.07	1.955	0.3220	57	-2.19	008.08	1
009	Fiber, Neutral Detergent (%)	0358	28.29	0.6700	30.19	1.420	0.4326	54	-1.34	009.09	0
009	Fiber, Neutral Detergent (%)	0049	28.41	0.1100	30.19	1.420	0.4326	54	-1.26	009.09	0
009	Fiber, Neutral Detergent (%)	0263	28.42	0.2248	30.19	1.420	0.4326	54	-1.25	009.09	0
009	Fiber, Neutral Detergent (%)	0598	28.45	0.1100	30.19	1.420	0.4326	54	-1.23	009.09	0
009	Fiber, Neutral Detergent (%)	0910	28.55	1.100	30.19	1.420	0.4326	54	-1.16	009.07	0
009	Fiber, Neutral Detergent (%)	0881	28.62	1.374	30.19	1.420	0.4326	54	-1.11	009.09	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
009	Fiber, Neutral Detergent (%)	0878	28.64	0.4700	30.19	1.420	0.4326	54	-1.10	009.09	0
009	Fiber, Neutral Detergent (%)	0294	28.70	0.0000	30.19	1.420	0.4326	54	-1.05	009.09	0
009	Fiber, Neutral Detergent (%)	0413	28.75	0.1000	30.19	1.420	0.4326	54	-1.02	009.09	0
009	Fiber, Neutral Detergent (%)	0354	28.82	0.0400	30.19	1.420	0.4326	54	-0.97	009.09	0
009	Fiber, Neutral Detergent (%)	0510	29.05	0.1000	30.19	1.420	0.4326	54	-0.80	009.09	0
009	Fiber, Neutral Detergent (%)	0948	29.06	0.6690	30.19	1.420	0.4326	54	-0.80	009.09	0
009	Fiber, Neutral Detergent (%)	0918	29.07	0.5307	30.19	1.420	0.4326	54	-0.79	009.09	0
009	Fiber, Neutral Detergent (%)	0164	29.10	0.2000	30.19	1.420	0.4326	54	-0.77	009.09	0
009	Fiber, Neutral Detergent (%)	0956	29.10	2.400	30.19	1.420	0.4326	54	-0.77	009.09	0
009	Fiber, Neutral Detergent (%)	0870	29.13	1.001	30.19	1.420	0.4326	54	-0.75	009.09	0
009	Fiber, Neutral Detergent (%)	0407	29.23	0.3109	30.19	1.420	0.4326	54	-0.68	009.09	0
009	Fiber, Neutral Detergent (%)	2196	29.45	0.0000	30.19	1.420	0.4326	54	-0.52	009.09	0
009	Fiber, Neutral Detergent (%)	0596	29.47	0.6500	30.19	1.420	0.4326	54	-0.51	009.09	0
009	Fiber, Neutral Detergent (%)	2012	29.53	0.2200	30.19	1.420	0.4326	54	-0.47	009.09	0
009	Fiber, Neutral Detergent (%)	0278	29.70	0.2000	30.19	1.420	0.4326	54	-0.35	009.09	0
009	Fiber, Neutral Detergent (%)	2193	29.84	0.1400	30.19	1.420	0.4326	54	-0.25	009.09	0
009	Fiber, Neutral Detergent (%)	0357	29.85	0.5000	30.19	1.420	0.4326	54	-0.24	009.09	0
009	Fiber, Neutral Detergent (%)	0948	29.98	0.4600	30.19	1.420	0.4326	54	-0.15	009.07	0
009	Fiber, Neutral Detergent (%)	0148	30.00	0.6000	30.19	1.420	0.4326	54	-0.14	009.09	0
009	Fiber, Neutral Detergent (%)	0581	30.02	0.4100	30.19	1.420	0.4326	54	-0.13	009.09	0
009	Fiber, Neutral Detergent (%)	0037	30.02	0.1200	30.19	1.420	0.4326	54	-0.12	009.09	0
009	Fiber, Neutral Detergent (%)	0939	30.07	0.3300	30.19	1.420	0.4326	54	-0.09	009.09	0
009	Fiber, Neutral Detergent (%)	0968	30.10	0.7140	30.19	1.420	0.4326	54	-0.07	009.09	0
009	Fiber, Neutral Detergent (%)	0912	30.20	0.4000	30.19	1.420	0.4326	54	0.01	009.09	0
009	Fiber, Neutral Detergent (%)	0265	30.25	0.5000	30.19	1.420	0.4326	54	0.04	009.09	0
009	Fiber, Neutral Detergent (%)	0353	30.39	0.7400	30.19	1.420	0.4326	54	0.14	009.07	0
009	Fiber, Neutral Detergent (%)	0675	30.45	0.1600	30.19	1.420	0.4326	54	0.18	009.09	0
009	Fiber, Neutral Detergent (%)	0675	30.51	0.1900	30.19	1.420	0.4326	54	0.22	009.07	0
009	Fiber, Neutral Detergent (%)	0083	30.53	0.1100	30.19	1.420	0.4326	54	0.23	009.09	0
009	Fiber, Neutral Detergent (%)	0889	30.71	0.2900	30.19	1.420	0.4326	54	0.36	009.99	0
009	Fiber, Neutral Detergent (%)	2192	30.75	0.0500	30.19	1.420	0.4326	54	0.39	009.07	0
009	Fiber, Neutral Detergent (%)	0345	30.90	0.6000	30.19	1.420	0.4326	54	0.50	009.07	0
009	Fiber, Neutral Detergent (%)	0590	30.90	0.4000	30.19	1.420	0.4326	54	0.50	009.09	0
009	Fiber, Neutral Detergent (%)	0098	31.00	0.6000	30.19	1.420	0.4326	54	0.57	009.07	0
009	Fiber, Neutral Detergent (%)	0693	31.05	0.2680	30.19	1.420	0.4326	54	0.61	009.09	0
009	Fiber, Neutral Detergent (%)	0848	31.24	0.0100	30.19	1.420	0.4326	54	0.73	009.09	0
009	Fiber, Neutral Detergent (%)	0066	31.26	0.2800	30.19	1.420	0.4326	54	0.75	009.09	0
009	Fiber, Neutral Detergent (%)	0226	31.48	0.5500	30.19	1.420	0.4326	54	0.90	009.07	0
009	Fiber, Neutral Detergent (%)	0676	31.67	0.6900	30.19	1.420	0.4326	54	1.04	009.07	0
009	Fiber, Neutral Detergent (%)	0646	31.77	1.600	30.19	1.420	0.4326	54	1.11	009.09	0
009	Fiber, Neutral Detergent (%)	0941	31.88	0.2200	30.19	1.420	0.4326	54	1.19	009.99	0
009	Fiber, Neutral Detergent (%)	0297	32.18	0.4800	30.19	1.420	0.4326	54	1.40	009.09	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
009	Fiber, Neutral Detergent (%)	0689	32.55	0.5000	30.19	1.420	0.4326	54	1.66	009.07	0
009	Fiber, Neutral Detergent (%)	0309	33.16	0.3101	30.19	1.420	0.4326	54	2.09	009.07	0
009	Fiber, Neutral Detergent (%)	0726	33.24	0.0100	30.19	1.420	0.4326	54	2.14	009.09	0
009	Fiber, Neutral Detergent (%)	0536	33.84	0.0100	30.19	1.420	0.4326	54	2.56	009.09	0
009	Fiber, Neutral Detergent (%)	0619	34.75	0.3000	30.19	1.420	0.4326	54	3.21	009.09	0
009	Fiber, Neutral Detergent (%)	0190	35.57	0.3400	30.19	1.420	0.4326	54	3.79	009.09	0
009	Fiber, Neutral Detergent (%)	0889	28.52	8.230	30.19	1.420	0.4326	54	-1.18	009.09	1
009	Fiber, Neutral Detergent (%)	2129	14.71	0.4350	30.19	1.420	0.4326	54	-10.90	009.09	2
010	Moisture (%)	0620	8.610	0.2663	9.573	0.5094	0.1048	16	-1.89	010.99	0
010	Moisture (%)	2089	9.090	0.2200	9.573	0.5094	0.1048	16	-0.95	010.99	0
010	Moisture (%)	0164	9.225	0.0300	9.573	0.5094	0.1048	16	-0.68	010.03	0
010	Moisture (%)	0726	9.225	0.0100	9.573	0.5094	0.1048	16	-0.68	010.99	0
010	Moisture (%)	0692	9.245	0.5700	9.573	0.5094	0.1048	16	-0.64	010.99	0
010	Moisture (%)	0948	9.344	0.0250	9.573	0.5094	0.1048	16	-0.45	010.99	0
010	Moisture (%)	2076	9.420	0.0200	9.573	0.5094	0.1048	16	-0.30	010.99	0
010	Moisture (%)	0939	9.450	0.1000	9.573	0.5094	0.1048	16	-0.24	010.99	0
010	Moisture (%)	0884	9.500	0.0000	9.573	0.5094	0.1048	16	-0.14	010.99	0
010	Moisture (%)	2150	9.565	0.1100	9.573	0.5094	0.1048	16	-0.02	010.99	0
010	Moisture (%)	0652	9.600	0.2000	9.573	0.5094	0.1048	16	0.05	010.99	0
010	Moisture (%)	0889	9.923	0.0050	9.573	0.5094	0.1048	16	0.69	010.11	0
010	Moisture (%)	0190	9.975	0.0100	9.573	0.5094	0.1048	16	0.79	010.99	0
010	Moisture (%)	0914	10.13	0.0400	9.573	0.5094	0.1048	16	1.09	010.99	0
010	Moisture (%)	0852	10.56	0.0500	9.573	0.5094	0.1048	16	1.93	010.11	0
010	Moisture (%)	0405	10.96	0.0200	9.573	0.5094	0.1048	16	2.72	010.99	0
010	Moisture (%)	0843	9.030	1.040	9.573	0.5094	0.1048	16	-1.07	010.03	1
010	Moisture (%)	2129	91.92	0.4670	9.573	0.5094	0.1048	16	161.67	010.11	3
010	Moisture (%)	2129	92.65	0.0350	9.573	0.5094	0.1048	16	163.09	010.99	3
011	Loss on Drying (%)	0520	8.775	0.0700	10.21	0.5013	0.1025	65	-2.87	011.01	0
011	Loss on Drying (%)	0643	8.900	0.2000	10.21	0.5013	0.1025	65	-2.62	011.01	0
011	Loss on Drying (%)	0574	9.085	0.1700	10.21	0.5013	0.1025	65	-2.25	011.01	0
011	Loss on Drying (%)	0175	9.250	0.1000	10.21	0.5013	0.1025	65	-1.92	011.01	0
011	Loss on Drying (%)	2200	9.250	0.1000	10.21	0.5013	0.1025	65	-1.92	011.01	0
011	Loss on Drying (%)	0682	9.260	0.0000	10.21	0.5013	0.1025	65	-1.90	011.01	0
011	Loss on Drying (%)	2304	9.330	0.2600	10.21	0.5013	0.1025	65	-1.76	011.99	0
011	Loss on Drying (%)	0650	9.410	0.5400	10.21	0.5013	0.1025	65	-1.60	011.01	0
011	Loss on Drying (%)	0968	9.475	0.0020	10.21	0.5013	0.1025	65	-1.47	011.01	0
011	Loss on Drying (%)	2199	9.500	0.2000	10.21	0.5013	0.1025	65	-1.42	011.01	0
011	Loss on Drying (%)	0685	9.625	0.0100	10.21	0.5013	0.1025	65	-1.17	011.01	0
011	Loss on Drying (%)	2033	9.705	0.0100	10.21	0.5013	0.1025	65	-1.01	011.01	0
011	Loss on Drying (%)	0596	9.735	0.3900	10.21	0.5013	0.1025	65	-0.95	011.01	0
011	Loss on Drying (%)	0164	9.795	0.1100	10.21	0.5013	0.1025	65	-0.83	011.01	0
011	Loss on Drying (%)	0622	9.800	0.3048	10.21	0.5013	0.1025	65	-0.83	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
011	Loss on Drying (%)	0536	9.940	0.0600	10.21	0.5013	0.1025	65	-0.55	011.01	0
011	Loss on Drying (%)	0563	9.954	0.0364	10.21	0.5013	0.1025	65	-0.52	011.01	0
011	Loss on Drying (%)	0294	10.00	0.0000	10.21	0.5013	0.1025	65	-0.43	011.01	0
011	Loss on Drying (%)	0675	10.03	0.0200	10.21	0.5013	0.1025	65	-0.37	011.01	0
011	Loss on Drying (%)	0425	10.05	0.1500	10.21	0.5013	0.1025	65	-0.34	011.01	0
011	Loss on Drying (%)	2181	10.05	0.2900	10.21	0.5013	0.1025	65	-0.34	011.01	0
011	Loss on Drying (%)	0618	10.06	0.0100	10.21	0.5013	0.1025	65	-0.32	011.01	0
011	Loss on Drying (%)	0620	10.10	0.1281	10.21	0.5013	0.1025	65	-0.23	011.01	0
011	Loss on Drying (%)	0226	10.14	0.0600	10.21	0.5013	0.1025	65	-0.15	011.01	0
011	Loss on Drying (%)	2168	10.15	0.0500	10.21	0.5013	0.1025	65	-0.14	011.01	0
011	Loss on Drying (%)	0881	10.15	0.1300	10.21	0.5013	0.1025	65	-0.12	011.01	0
011	Loss on Drying (%)	0511	10.22	0.0900	10.21	0.5013	0.1025	65	0.00	011.01	0
011	Loss on Drying (%)	0953	10.22	0.0300	10.21	0.5013	0.1025	65	0.00	011.01	0
011	Loss on Drying (%)	2006	10.23	0.0400	10.21	0.5013	0.1025	65	0.03	011.01	0
011	Loss on Drying (%)	0859	10.24	0.1170	10.21	0.5013	0.1025	65	0.05	011.01	0
011	Loss on Drying (%)	0598	10.25	0.0300	10.21	0.5013	0.1025	65	0.06	011.01	0
011	Loss on Drying (%)	0047	10.25	0.1000	10.21	0.5013	0.1025	65	0.07	011.01	0
011	Loss on Drying (%)	0674	10.26	0.0000	10.21	0.5013	0.1025	65	0.09	011.01	0
011	Loss on Drying (%)	0062	10.27	0.1100	10.21	0.5013	0.1025	65	0.12	011.01	0
011	Loss on Drying (%)	0100	10.29	0.0900	10.21	0.5013	0.1025	65	0.14	011.01	0
011	Loss on Drying (%)	2196	10.33	0.0000	10.21	0.5013	0.1025	65	0.23	011.01	0
011	Loss on Drying (%)	0852	10.35	0.0800	10.21	0.5013	0.1025	65	0.27	011.99	0
011	Loss on Drying (%)	0653	10.36	0.1400	10.21	0.5013	0.1025	65	0.29	011.01	0
011	Loss on Drying (%)	2174	10.37	0.0700	10.21	0.5013	0.1025	65	0.30	011.01	0
011	Loss on Drying (%)	0144	10.39	0.3700	10.21	0.5013	0.1025	65	0.34	011.01	0
011	Loss on Drying (%)	0553	10.39	0.3500	10.21	0.5013	0.1025	65	0.34	011.01	0
011	Loss on Drying (%)	0354	10.40	0.0100	10.21	0.5013	0.1025	65	0.36	011.01	0
011	Loss on Drying (%)	0171	10.40	0.0800	10.21	0.5013	0.1025	65	0.37	011.01	0
011	Loss on Drying (%)	2073	10.43	0.0200	10.21	0.5013	0.1025	65	0.43	011.01	0
011	Loss on Drying (%)	0870	10.43	0.0716	10.21	0.5013	0.1025	65	0.44	011.01	0
011	Loss on Drying (%)	0263	10.45	0.0700	10.21	0.5013	0.1025	65	0.46	011.01	0
011	Loss on Drying (%)	0098	10.51	0.0300	10.21	0.5013	0.1025	65	0.58	011.01	0
011	Loss on Drying (%)	0539	10.52	0.0800	10.21	0.5013	0.1025	65	0.61	011.01	0
011	Loss on Drying (%)	0123	10.56	0.0100	10.21	0.5013	0.1025	65	0.68	011.01	0
011	Loss on Drying (%)	0242	10.56	0.1000	10.21	0.5013	0.1025	65	0.69	011.01	0
011	Loss on Drying (%)	0529	10.58	0.0100	10.21	0.5013	0.1025	65	0.72	011.02	0
011	Loss on Drying (%)	0229	10.58	0.0040	10.21	0.5013	0.1025	65	0.73	011.01	0
011	Loss on Drying (%)	2193	10.60	0.0700	10.21	0.5013	0.1025	65	0.76	011.01	0
011	Loss on Drying (%)	0066	10.62	0.0400	10.21	0.5013	0.1025	65	0.81	011.01	0
011	Loss on Drying (%)	0510	10.65	0.1000	10.21	0.5013	0.1025	65	0.87	011.01	0
011	Loss on Drying (%)	0541	10.69	0.2800	10.21	0.5013	0.1025	65	0.95	011.01	0
011	Loss on Drying (%)	0843	10.71	0.2000	10.21	0.5013	0.1025	65	0.99	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
011	Loss on Drying (%)	0407	10.76	0.0607	10.21	0.5013	0.1025	65	1.08	011.01	0
011	Loss on Drying (%)	2022	10.78	0.0600	10.21	0.5013	0.1025	65	1.13	011.01	0
011	Loss on Drying (%)	0589	10.80	0.0000	10.21	0.5013	0.1025	65	1.17	011.01	0
011	Loss on Drying (%)	0309	10.86	0.1500	10.21	0.5013	0.1025	65	1.28	011.01	0
011	Loss on Drying (%)	0358	10.88	0.1400	10.21	0.5013	0.1025	65	1.33	011.01	0
011	Loss on Drying (%)	0202	10.92	0.0300	10.21	0.5013	0.1025	65	1.40	011.01	0
011	Loss on Drying (%)	0004	10.94	0.0100	10.21	0.5013	0.1025	65	1.44	011.01	0
011	Loss on Drying (%)	0298	12.47	0.0200	10.21	0.5013	0.1025	65	4.50	011.01	0
011	Loss on Drying (%)	0942	9.985	0.7300	10.21	0.5013	0.1025	65	-0.46	011.02	1
011	Loss on Drying (%)	0265	10.35	0.7000	10.21	0.5013	0.1025	65	0.27	011.01	1
012	Starch (%)	0910	13.45	0.3000	19.08	1.835	0.3980	38	-3.07	012.01	0
012	Starch (%)	0510	15.20	0.0000	19.08	1.835	0.3980	38	-2.11	012.04	0
012	Starch (%)	0676	15.90	0.5600	19.08	1.835	0.3980	38	-1.73	012.01	0
012	Starch (%)	0870	16.58	0.3543	19.08	1.835	0.3980	38	-1.36	012.01	0
012	Starch (%)	0164	17.15	1.500	19.08	1.835	0.3980	38	-1.05	012.01	0
012	Starch (%)	0956	17.35	0.3000	19.08	1.835	0.3980	38	-0.94	012.01	0
012	Starch (%)	0098	17.50	0.4000	19.08	1.835	0.3980	38	-0.86	012.04	0
012	Starch (%)	2181	18.10	0.1100	19.08	1.835	0.3980	38	-0.54	012.00	0
012	Starch (%)	0297	18.12	1.240	19.08	1.835	0.3980	38	-0.52	012.03	0
012	Starch (%)	0164	18.25	0.7000	19.08	1.835	0.3980	38	-0.45	012.00	0
012	Starch (%)	0148	18.25	0.1000	19.08	1.835	0.3980	38	-0.45	012.01	0
012	Starch (%)	2009	18.40	0.0466	19.08	1.835	0.3980	38	-0.37	012.00	0
012	Starch (%)	0353	18.40	0.2800	19.08	1.835	0.3980	38	-0.37	012.20	0
012	Starch (%)	0693	18.58	0.2745	19.08	1.835	0.3980	38	-0.27	012.01	0
012	Starch (%)	2129	18.63	1.053	19.08	1.835	0.3980	38	-0.24	012.11	0
012	Starch (%)	2193	18.80	0.1200	19.08	1.835	0.3980	38	-0.15	012.00	0
012	Starch (%)	0278	18.80	0.6000	19.08	1.835	0.3980	38	-0.15	012.04	0
012	Starch (%)	0689	18.85	0.3000	19.08	1.835	0.3980	38	-0.12	012.00	0
012	Starch (%)	0354	18.89	0.4300	19.08	1.835	0.3980	38	-0.11	012.00	0
012	Starch (%)	0619	18.90	0.2000	19.08	1.835	0.3980	38	-0.10	012.00	0
012	Starch (%)	0910	18.95	0.7000	19.08	1.835	0.3980	38	-0.07	012.00	0
012	Starch (%)	0653	19.04	0.2000	19.08	1.835	0.3980	38	-0.02	012.00	0
012	Starch (%)	2006	19.15	0.5400	19.08	1.835	0.3980	38	0.04	012.00	0
012	Starch (%)	2196	19.23	0.0000	19.08	1.835	0.3980	38	0.08	012.00	0
012	Starch (%)	0948	19.23	0.3600	19.08	1.835	0.3980	38	0.08	012.04	0
012	Starch (%)	0941	19.24	0.3200	19.08	1.835	0.3980	38	0.09	012.00	0
012	Starch (%)	0610	19.25	0.1000	19.08	1.835	0.3980	38	0.09	012.00	0
012	Starch (%)	0878	19.55	0.5000	19.08	1.835	0.3980	38	0.26	012.00	0
012	Starch (%)	0066	19.97	0.6300	19.08	1.835	0.3980	38	0.48	012.03	0
012	Starch (%)	0265	20.10	0.0000	19.08	1.835	0.3980	38	0.56	012.01	0
012	Starch (%)	0889	21.01	0.0200	19.08	1.835	0.3980	38	1.05	012.11	0
012	Starch (%)	0852	21.18	0.2200	19.08	1.835	0.3980	38	1.14	012.11	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
012	Starch (%)	2089	21.45	0.4500	19.08	1.835	0.3980	38	1.29	012.01	0
012	Starch (%)	2192	21.96	0.1300	19.08	1.835	0.3980	38	1.57	012.00	0
012	Starch (%)	0918	23.24	1.314	19.08	1.835	0.3980	38	2.27	012.01	0
012	Starch (%)	0297	23.89	0.2200	19.08	1.835	0.3980	38	2.62	012.11	0
012	Starch (%)	2129	26.36	0.4910	19.08	1.835	0.3980	38	3.97	012.04	0
012	Starch (%)	0889	34.30	0.0600	19.08	1.835	0.3980	38	8.29	012.03	0
012	Starch (%)	0407	14.51	2.026	19.08	1.835	0.3980	38	-2.49	012.03	1
013	Fat, Acid Pretreat (%)	0618	3.695	0.1276	4.955	0.6594	0.2178	47	-1.91	013.08	0
013	Fat, Acid Pretreat (%)	0096	3.742	0.1687	4.955	0.6594	0.2178	47	-1.84	013.00	0
013	Fat, Acid Pretreat (%)	0083	4.005	0.0300	4.955	0.6594	0.2178	47	-1.44	013.00	0
013	Fat, Acid Pretreat (%)	2146	4.080	0.2000	4.955	0.6594	0.2178	47	-1.33	013.00	0
013	Fat, Acid Pretreat (%)	0541	4.085	0.2700	4.955	0.6594	0.2178	47	-1.32	013.00	0
013	Fat, Acid Pretreat (%)	0689	4.200	0.0000	4.955	0.6594	0.2178	47	-1.15	013.00	0
013	Fat, Acid Pretreat (%)	0036	4.200	0.1057	4.955	0.6594	0.2178	47	-1.15	013.10	0
013	Fat, Acid Pretreat (%)	2073	4.240	0.3800	4.955	0.6594	0.2178	47	-1.09	013.00	0
013	Fat, Acid Pretreat (%)	0242	4.250	0.0400	4.955	0.6594	0.2178	47	-1.07	013.00	0
013	Fat, Acid Pretreat (%)	0939	4.250	0.1800	4.955	0.6594	0.2178	47	-1.07	013.10	0
013	Fat, Acid Pretreat (%)	0853	4.400	0.0000	4.955	0.6594	0.2178	47	-0.84	013.02	0
013	Fat, Acid Pretreat (%)	0968	4.530	0.0030	4.955	0.6594	0.2178	47	-0.65	013.13	0
013	Fat, Acid Pretreat (%)	0407	4.533	0.2476	4.955	0.6594	0.2178	47	-0.64	013.13	0
013	Fat, Acid Pretreat (%)	0297	4.650	0.2400	4.955	0.6594	0.2178	47	-0.46	013.00	0
013	Fat, Acid Pretreat (%)	0229	4.670	0.3600	4.955	0.6594	0.2178	47	-0.43	013.00	0
013	Fat, Acid Pretreat (%)	2196	4.670	0.0000	4.955	0.6594	0.2178	47	-0.43	013.10	0
013	Fat, Acid Pretreat (%)	0353	4.690	0.1400	4.955	0.6594	0.2178	47	-0.40	013.10	0
013	Fat, Acid Pretreat (%)	0650	4.720	0.1000	4.955	0.6594	0.2178	47	-0.36	013.02	0
013	Fat, Acid Pretreat (%)	0309	4.837	0.1608	4.955	0.6594	0.2178	47	-0.18	013.00	0
013	Fat, Acid Pretreat (%)	0870	4.843	0.1019	4.955	0.6594	0.2178	47	-0.17	013.02	0
013	Fat, Acid Pretreat (%)	0921	4.920	0.1800	4.955	0.6594	0.2178	47	-0.05	013.02	0
013	Fat, Acid Pretreat (%)	0843	4.925	0.5300	4.955	0.6594	0.2178	47	-0.05	013.13	0
013	Fat, Acid Pretreat (%)	0910	4.970	0.2800	4.955	0.6594	0.2178	47	0.02	013.00	0
013	Fat, Acid Pretreat (%)	0914	4.995	0.1700	4.955	0.6594	0.2178	47	0.06	013.02	0
013	Fat, Acid Pretreat (%)	0652	5.000	0.0000	4.955	0.6594	0.2178	47	0.07	013.00	0
013	Fat, Acid Pretreat (%)	0884	5.000	0.4000	4.955	0.6594	0.2178	47	0.07	013.00	0
013	Fat, Acid Pretreat (%)	0948	5.025	0.0700	4.955	0.6594	0.2178	47	0.11	013.00	0
013	Fat, Acid Pretreat (%)	2304	5.055	0.9700	4.955	0.6594	0.2178	47	0.15	013.99	0
013	Fat, Acid Pretreat (%)	0098	5.125	0.0900	4.955	0.6594	0.2178	47	0.26	013.02	0
013	Fat, Acid Pretreat (%)	0354	5.155	0.2900	4.955	0.6594	0.2178	47	0.30	013.02	0
013	Fat, Acid Pretreat (%)	0682	5.240	0.0000	4.955	0.6594	0.2178	47	0.43	013.02	0
013	Fat, Acid Pretreat (%)	0100	5.340	0.2800	4.955	0.6594	0.2178	47	0.58	013.02	0
013	Fat, Acid Pretreat (%)	0407	5.371	0.6420	4.955	0.6594	0.2178	47	0.63	013.02	0
013	Fat, Acid Pretreat (%)	0675	5.445	0.0500	4.955	0.6594	0.2178	47	0.74	013.02	0
013	Fat, Acid Pretreat (%)	0148	5.465	0.1700	4.955	0.6594	0.2178	47	0.77	013.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
013	Fat, Acid Pretreat (%)	2150	5.480	0.0800	4.955	0.6594	0.2178	47	0.80	013.00	0
013	Fat, Acid Pretreat (%)	0581	5.540	0.3800	4.955	0.6594	0.2178	47	0.89	013.13	0
013	Fat, Acid Pretreat (%)	0171	5.555	0.1500	4.955	0.6594	0.2178	47	0.91	013.02	0
013	Fat, Acid Pretreat (%)	2259	5.568	0.9310	4.955	0.6594	0.2178	47	0.93	013.02	0
013	Fat, Acid Pretreat (%)	0676	5.570	0.2600	4.955	0.6594	0.2178	47	0.93	013.02	0
013	Fat, Acid Pretreat (%)	0553	5.605	0.2100	4.955	0.6594	0.2178	47	0.99	013.02	0
013	Fat, Acid Pretreat (%)	0010	5.650	0.1000	4.955	0.6594	0.2178	47	1.05	013.02	0
013	Fat, Acid Pretreat (%)	2076	5.718	0.5386	4.955	0.6594	0.2178	47	1.16	013.00	0
013	Fat, Acid Pretreat (%)	0139	5.750	0.1800	4.955	0.6594	0.2178	47	1.20	013.00	0
013	Fat, Acid Pretreat (%)	0164	5.785	0.0300	4.955	0.6594	0.2178	47	1.26	013.02	0
013	Fat, Acid Pretreat (%)	0202	5.925	0.2700	4.955	0.6594	0.2178	47	1.47	013.00	0
013	Fat, Acid Pretreat (%)	0265	6.715	0.1300	4.955	0.6594	0.2178	47	2.67	013.13	0
013	Fat, Acid Pretreat (%)	0027	5.930	1.818	4.955	0.6594	0.2178	47	1.48	013.13	1
015	Aluminum (ppm)	2129	49.13	12.52	169.6	28.62	9.091	18	-4.21	015.42	0
015	Aluminum (ppm)	0037	136.8	9.800	169.6	28.62	9.091	18	-1.15	015.42	0
015	Aluminum (ppm)	0171	138.0	1.100	169.6	28.62	9.091	18	-1.11	015.41	0
015	Aluminum (ppm)	0049	143.7	12.45	169.6	28.62	9.091	18	-0.91	015.41	0
015	Aluminum (ppm)	0560	149.4	2.900	169.6	28.62	9.091	18	-0.71	015.52	0
015	Aluminum (ppm)	0169	149.5	5.000	169.6	28.62	9.091	18	-0.70	015.43	0
015	Aluminum (ppm)	0297	159.0	26.00	169.6	28.62	9.091	18	-0.37	015.43	0
015	Aluminum (ppm)	0510	163.0	2.000	169.6	28.62	9.091	18	-0.23	015.43	0
015	Aluminum (ppm)	0918	170.1	1.022	169.6	28.62	9.091	18	0.02	015.43	0
015	Aluminum (ppm)	0164	175.5	13.00	169.6	28.62	9.091	18	0.21	015.41	0
015	Aluminum (ppm)	0553	181.5	9.000	169.6	28.62	9.091	18	0.41	015.53	0
015	Aluminum (ppm)	0345	183.4	3.700	169.6	28.62	9.091	18	0.48	015.43	0
015	Aluminum (ppm)	0227	184.0	4.000	169.6	28.62	9.091	18	0.50	015.41	0
015	Aluminum (ppm)	0939	187.5	1.000	169.6	28.62	9.091	18	0.62	015.53	0
015	Aluminum (ppm)	0407	187.7	30.83	169.6	28.62	9.091	18	0.63	015.41	0
015	Aluminum (ppm)	0407	199.2	8.311	169.6	28.62	9.091	18	1.03	015.43	0
015	Aluminum (ppm)	0353	206.0	10.00	169.6	28.62	9.091	18	1.27	015.43	0
015	Aluminum (ppm)	0520	215.5	11.00	169.6	28.62	9.091	18	1.60	015.41	0
015	Aluminum (ppm)	0425	134.8	66.30	169.6	28.62	9.091	18	-1.22	015.43	1
017	Boron (ppm)	0229	0.5000	0.0000	4.221	1.079	0.5334	18	-3.45	017.41	0
017	Boron (ppm)	0553	3.230	0.6200	4.221	1.079	0.5334	18	-0.92	017.53	0
017	Boron (ppm)	0171	3.250	0.3000	4.221	1.079	0.5334	18	-0.90	017.41	0
017	Boron (ppm)	0358	3.275	0.0700	4.221	1.079	0.5334	18	-0.88	017.41	0
017	Boron (ppm)	0693	3.639	0.2140	4.221	1.079	0.5334	18	-0.54	017.42	0
017	Boron (ppm)	0037	3.700	0.2000	4.221	1.079	0.5334	18	-0.48	017.42	0
017	Boron (ppm)	0560	3.775	0.7900	4.221	1.079	0.5334	18	-0.41	017.52	0
017	Boron (ppm)	0265	3.885	0.1900	4.221	1.079	0.5334	18	-0.31	017.42	0
017	Boron (ppm)	0049	3.980	1.380	4.221	1.079	0.5334	18	-0.22	017.41	0
017	Boron (ppm)	0510	4.000	0.0000	4.221	1.079	0.5334	18	-0.20	017.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
017	Boron (ppm)	0345	4.250	0.3000	4.221	1.079	0.5334	18	0.03	017.43	0
017	Boron (ppm)	0083	4.300	0.4000	4.221	1.079	0.5334	18	0.07	017.43	0
017	Boron (ppm)	0918	4.535	0.0500	4.221	1.079	0.5334	18	0.29	017.43	0
017	Boron (ppm)	0226	5.120	0.2200	4.221	1.079	0.5334	18	0.83	017.41	0
017	Boron (ppm)	0407	5.173	2.246	4.221	1.079	0.5334	18	0.88	017.41	0
017	Boron (ppm)	0407	5.585	2.361	4.221	1.079	0.5334	18	1.26	017.43	0
017	Boron (ppm)	0294	6.720	0.0400	4.221	1.079	0.5334	18	2.32	017.42	0
017	Boron (ppm)	2129	10.15	0.2200	4.221	1.079	0.5334	18	5.50	017.42	0
017	Boron (ppm)	0353	7.000	4.000	4.221	1.079	0.5334	18	2.58	017.43	1
017	Boron (ppm)	0297	0.0000	0.0000	4.221	1.079	0.5334	18		017.43	4
019	Calcium (%)	0875	1.720	0.0600	2.179	0.1272	0.0611	131	-3.60	019.43	0
019	Calcium (%)	0889	1.750	0.0600	2.179	0.1272	0.0611	131	-3.37	019.99	0
019	Calcium (%)	2200	1.800	0.2000	2.179	0.1272	0.0611	131	-2.97	019.08	0
019	Calcium (%)	2199	1.920	0.1600	2.179	0.1272	0.0611	131	-2.03	019.08	0
019	Calcium (%)	0154	1.935	0.0519	2.179	0.1272	0.0611	131	-1.92	019.52	0
019	Calcium (%)	0144	1.945	0.0900	2.179	0.1272	0.0611	131	-1.84	019.41	0
019	Calcium (%)	0038	1.964	0.0210	2.179	0.1272	0.0611	131	-1.69	019.43	0
019	Calcium (%)	2012	1.975	0.0100	2.179	0.1272	0.0611	131	-1.60	019.41	0
019	Calcium (%)	0175	1.995	0.1300	2.179	0.1272	0.0611	131	-1.44	019.31	0
019	Calcium (%)	0298	2.000	0.0200	2.179	0.1272	0.0611	131	-1.40	019.41	0
019	Calcium (%)	0939	2.005	0.0300	2.179	0.1272	0.0611	131	-1.36	019.31	0
019	Calcium (%)	0682	2.010	0.0000	2.179	0.1272	0.0611	131	-1.32	019.43	0
019	Calcium (%)	0939	2.015	0.0500	2.179	0.1272	0.0611	131	-1.29	019.53	0
019	Calcium (%)	0357	2.025	0.0074	2.179	0.1272	0.0611	131	-1.21	019.42	0
019	Calcium (%)	0038	2.039	0.1003	2.179	0.1272	0.0611	131	-1.10	019.31	0
019	Calcium (%)	0884	2.045	0.2100	2.179	0.1272	0.0611	131	-1.05	019.31	0
019	Calcium (%)	0242	2.045	0.1100	2.179	0.1272	0.0611	131	-1.05	019.99	0
019	Calcium (%)	0918	2.048	0.0750	2.179	0.1272	0.0611	131	-1.03	019.43	0
019	Calcium (%)	0693	2.049	0.0808	2.179	0.1272	0.0611	131	-1.02	019.42	0
019	Calcium (%)	0536	2.060	0.1000	2.179	0.1272	0.0611	131	-0.93	019.02	0
019	Calcium (%)	0675	2.060	0.0600	2.179	0.1272	0.0611	131	-0.93	019.43	0
019	Calcium (%)	0049	2.070	0.2000	2.179	0.1272	0.0611	131	-0.85	019.41	0
019	Calcium (%)	0171	2.070	0.0000	2.179	0.1272	0.0611	131	-0.85	019.41	0
019	Calcium (%)	0610	2.070	0.0000	2.179	0.1272	0.0611	131	-0.85	019.43	0
019	Calcium (%)	2181	2.075	0.0500	2.179	0.1272	0.0611	131	-0.81	019.00	0
019	Calcium (%)	0123	2.075	0.1100	2.179	0.1272	0.0611	131	-0.81	019.41	0
019	Calcium (%)	0685	2.075	0.0100	2.179	0.1272	0.0611	131	-0.81	019.41	0
019	Calcium (%)	0229	2.078	0.0480	2.179	0.1272	0.0611	131	-0.79	019.41	0
019	Calcium (%)	0004	2.080	0.0800	2.179	0.1272	0.0611	131	-0.77	019.41	0
019	Calcium (%)	2295	2.085	0.0100	2.179	0.1272	0.0611	131	-0.74	019.00	0
019	Calcium (%)	0003	2.090	0.0600	2.179	0.1272	0.0611	131	-0.70	019.41	0
019	Calcium (%)	2062	2.091	0.0093	2.179	0.1272	0.0611	131	-0.69	019.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
019	Calcium (%)	0620	2.096	0.1169	2.179	0.1272	0.0611	131	-0.65	019.31	0
019	Calcium (%)	0921	2.097	0.0016	2.179	0.1272	0.0611	131	-0.64	019.31	0
019	Calcium (%)	0674	2.100	0.0000	2.179	0.1272	0.0611	131	-0.62	019.31	0
019	Calcium (%)	0033	2.100	0.1600	2.179	0.1272	0.0611	131	-0.62	019.43	0
019	Calcium (%)	2109	2.106	0.0560	2.179	0.1272	0.0611	131	-0.57	019.00	0
019	Calcium (%)	0968	2.107	0.0080	2.179	0.1272	0.0611	131	-0.56	019.43	0
019	Calcium (%)	0563	2.108	0.0740	2.179	0.1272	0.0611	131	-0.55	019.31	0
019	Calcium (%)	0676	2.110	0.0400	2.179	0.1272	0.0611	131	-0.54	019.99	0
019	Calcium (%)	0098	2.110	0.0600	2.179	0.1272	0.0611	131	-0.54	019.41	0
019	Calcium (%)	0036	2.117	0.1158	2.179	0.1272	0.0611	131	-0.49	019.43	0
019	Calcium (%)	0265	2.120	0.0000	2.179	0.1272	0.0611	131	-0.46	019.42	0
019	Calcium (%)	0590	2.125	0.0100	2.179	0.1272	0.0611	131	-0.42	019.99	0
019	Calcium (%)	0910	2.130	0.1200	2.179	0.1272	0.0611	131	-0.38	019.41	0
019	Calcium (%)	0297	2.130	0.0200	2.179	0.1272	0.0611	131	-0.38	019.43	0
019	Calcium (%)	0074	2.135	0.1100	2.179	0.1272	0.0611	131	-0.34	019.41	0
019	Calcium (%)	0009	2.135	0.1300	2.179	0.1272	0.0611	131	-0.34	019.42	0
019	Calcium (%)	2259	2.136	0.0370	2.179	0.1272	0.0611	131	-0.34	019.41	0
019	Calcium (%)	2006	2.137	0.0380	2.179	0.1272	0.0611	131	-0.33	019.09	0
019	Calcium (%)	0868	2.140	0.0200	2.179	0.1272	0.0611	131	-0.30	019.31	0
019	Calcium (%)	0083	2.140	0.0400	2.179	0.1272	0.0611	131	-0.30	019.43	0
019	Calcium (%)	0893	2.145	0.0100	2.179	0.1272	0.0611	131	-0.26	019.00	0
019	Calcium (%)	2146	2.145	0.0300	2.179	0.1272	0.0611	131	-0.26	019.31	0
019	Calcium (%)	2006	2.149	0.0290	2.179	0.1272	0.0611	131	-0.24	019.00	0
019	Calcium (%)	2022	2.150	0.0800	2.179	0.1272	0.0611	131	-0.22	019.00	0
019	Calcium (%)	0852	2.150	0.1000	2.179	0.1272	0.0611	131	-0.22	019.08	0
019	Calcium (%)	0687	2.150	0.1000	2.179	0.1272	0.0611	131	-0.22	019.31	0
019	Calcium (%)	0948	2.157	0.0300	2.179	0.1272	0.0611	131	-0.17	019.31	0
019	Calcium (%)	0358	2.160	0.1000	2.179	0.1272	0.0611	131	-0.15	019.41	0
019	Calcium (%)	0878	2.160	0.0800	2.179	0.1272	0.0611	131	-0.15	019.41	0
019	Calcium (%)	0505	2.160	0.0400	2.179	0.1272	0.0611	131	-0.15	019.43	0
019	Calcium (%)	0878	2.165	0.0100	2.179	0.1272	0.0611	131	-0.11	019.00	0
019	Calcium (%)	0010	2.165	0.0100	2.179	0.1272	0.0611	131	-0.11	019.43	0
019	Calcium (%)	0622	2.165	0.0105	2.179	0.1272	0.0611	131	-0.10	019.00	0
019	Calcium (%)	0407	2.167	0.0076	2.179	0.1272	0.0611	131	-0.09	019.41	0
019	Calcium (%)	0914	2.174	0.2501	2.179	0.1272	0.0611	131	-0.03	019.43	0
019	Calcium (%)	0590	2.175	0.0100	2.179	0.1272	0.0611	131	-0.03	019.08	0
019	Calcium (%)	0019	2.180	0.0600	2.179	0.1272	0.0611	131	0.01	019.41	0
019	Calcium (%)	0169	2.180	0.0400	2.179	0.1272	0.0611	131	0.01	019.32	0
019	Calcium (%)	0164	2.185	0.1500	2.179	0.1272	0.0611	131	0.05	019.44	0
019	Calcium (%)	2196	2.186	0.0000	2.179	0.1272	0.0611	131	0.06	019.08	0
019	Calcium (%)	0199	2.190	0.1600	2.179	0.1272	0.0611	131	0.09	019.53	0
019	Calcium (%)	0553	2.190	0.0000	2.179	0.1272	0.0611	131	0.09	019.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
019	Calcium (%)	0653	2.192	0.0080	2.179	0.1272	0.0611	131	0.11	019.41	0
019	Calcium (%)	0572	2.195	0.1300	2.179	0.1272	0.0611	131	0.13	019.53	0
019	Calcium (%)	0897	2.195	0.0100	2.179	0.1272	0.0611	131	0.13	019.00	0
019	Calcium (%)	0139	2.195	0.0300	2.179	0.1272	0.0611	131	0.13	019.41	0
019	Calcium (%)	0425	2.195	0.0100	2.179	0.1272	0.0611	131	0.13	019.43	0
019	Calcium (%)	0066	2.196	0.0580	2.179	0.1272	0.0611	131	0.14	019.43	0
019	Calcium (%)	2193	2.200	0.0200	2.179	0.1272	0.0611	131	0.17	019.08	0
019	Calcium (%)	2196	2.200	0.0000	2.179	0.1272	0.0611	131	0.17	019.31	0
019	Calcium (%)	2009	2.208	0.0053	2.179	0.1272	0.0611	131	0.23	019.08	0
019	Calcium (%)	0142	2.210	0.0974	2.179	0.1272	0.0611	131	0.24	019.31	0
019	Calcium (%)	2146	2.210	0.1000	2.179	0.1272	0.0611	131	0.25	019.51	0
019	Calcium (%)	0848	2.215	0.0500	2.179	0.1272	0.0611	131	0.29	019.41	0
019	Calcium (%)	0536	2.225	0.0100	2.179	0.1272	0.0611	131	0.37	019.31	0
019	Calcium (%)	0035	2.225	0.0900	2.179	0.1272	0.0611	131	0.37	019.42	0
019	Calcium (%)	0692	2.230	0.3000	2.179	0.1272	0.0611	131	0.40	019.42	0
019	Calcium (%)	0870	2.233	0.1210	2.179	0.1272	0.0611	131	0.42	019.42	0
019	Calcium (%)	0529	2.235	0.0700	2.179	0.1272	0.0611	131	0.44	019.31	0
019	Calcium (%)	0354	2.235	0.0100	2.179	0.1272	0.0611	131	0.44	019.41	0
019	Calcium (%)	0510	2.235	0.0100	2.179	0.1272	0.0611	131	0.44	019.43	0
019	Calcium (%)	0017	2.246	0.0063	2.179	0.1272	0.0611	131	0.53	019.43	0
019	Calcium (%)	2144	2.255	0.0300	2.179	0.1272	0.0611	131	0.60	019.00	0
019	Calcium (%)	0619	2.255	0.0300	2.179	0.1272	0.0611	131	0.60	019.41	0
019	Calcium (%)	0263	2.258	0.0537	2.179	0.1272	0.0611	131	0.62	019.42	0
019	Calcium (%)	2152	2.260	0.0200	2.179	0.1272	0.0611	131	0.64	019.08	0
019	Calcium (%)	2089	2.265	0.0700	2.179	0.1272	0.0611	131	0.68	019.43	0
019	Calcium (%)	0345	2.267	0.0130	2.179	0.1272	0.0611	131	0.69	019.43	0
019	Calcium (%)	0006	2.269	0.0220	2.179	0.1272	0.0611	131	0.71	019.43	0
019	Calcium (%)	0278	2.270	0.0200	2.179	0.1272	0.0611	131	0.72	019.42	0
019	Calcium (%)	0413	2.285	0.0900	2.179	0.1272	0.0611	131	0.84	019.41	0
019	Calcium (%)	2152	2.285	0.0100	2.179	0.1272	0.0611	131	0.84	019.44	0
019	Calcium (%)	0560	2.292	0.1996	2.179	0.1272	0.0611	131	0.89	019.52	0
019	Calcium (%)	2053	2.300	0.0000	2.179	0.1272	0.0611	131	0.95	019.42	0
019	Calcium (%)	2192	2.300	0.1600	2.179	0.1272	0.0611	131	0.95	019.43	0
019	Calcium (%)	0035	2.310	0.0200	2.179	0.1272	0.0611	131	1.03	019.43	0
019	Calcium (%)	0148	2.315	0.0900	2.179	0.1272	0.0611	131	1.07	019.41	0
019	Calcium (%)	0100	2.315	0.1500	2.179	0.1272	0.0611	131	1.07	019.99	0
019	Calcium (%)	0027	2.321	0.0850	2.179	0.1272	0.0611	131	1.12	019.43	0
019	Calcium (%)	0881	2.322	0.0580	2.179	0.1272	0.0611	131	1.13	019.00	0
019	Calcium (%)	0098	2.325	0.0500	2.179	0.1272	0.0611	131	1.15	019.43	0
019	Calcium (%)	2246	2.330	0.0200	2.179	0.1272	0.0611	131	1.19	019.31	0
019	Calcium (%)	0202	2.334	0.0668	2.179	0.1272	0.0611	131	1.22	019.43	0
019	Calcium (%)	0089	2.345	0.0100	2.179	0.1272	0.0611	131	1.31	019.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	0001	2.347	0.1050	2.179	0.1272	0.0611	131	1.32	019.31	0
019	Calcium (%)	0598	2.349	0.0760	2.179	0.1272	0.0611	131	1.34	019.41	0
019	Calcium (%)	0353	2.355	0.0900	2.179	0.1272	0.0611	131	1.39	019.43	0
019	Calcium (%)	0512	2.356	0.0450	2.179	0.1272	0.0611	131	1.39	019.43	0
019	Calcium (%)	0036	2.359	0.0712	2.179	0.1272	0.0611	131	1.42	019.03	0
019	Calcium (%)	0226	2.365	0.0300	2.179	0.1272	0.0611	131	1.47	019.41	0
019	Calcium (%)	0037	2.378	0.0250	2.179	0.1272	0.0611	131	1.56	019.42	0
019	Calcium (%)	0190	2.380	0.0400	2.179	0.1272	0.0611	131	1.58	019.42	0
019	Calcium (%)	0407	2.402	0.0489	2.179	0.1272	0.0611	131	1.76	019.43	0
019	Calcium (%)	0622	2.412	0.1058	2.179	0.1272	0.0611	131	1.83	019.31	0
019	Calcium (%)	0047	2.415	0.2100	2.179	0.1272	0.0611	131	1.86	019.52	0
019	Calcium (%)	0294	2.460	0.0400	2.179	0.1272	0.0611	131	2.21	019.42	0
019	Calcium (%)	0726	2.525	0.0100	2.179	0.1272	0.0611	131	2.72	019.42	0
019	Calcium (%)	0032	2.615	0.0300	2.179	0.1272	0.0611	131	3.43	019.43	0
019	Calcium (%)	0520	3.082	0.0070	2.179	0.1272	0.0611	131	7.10	019.41	0
019	Calcium (%)	0042	2.000	0.3200	2.179	0.1272	0.0611	131	-1.40	019.42	1
019	Calcium (%)	0018	2.120	0.3600	2.179	0.1272	0.0611	131	-0.46	019.41	1
019	Calcium (%)	0366	2.270	0.3600	2.179	0.1272	0.0611	131	0.72	019.42	1
019	Calcium (%)	2129	1.093	0.0200	2.179	0.1272	0.0611	131	-8.53	019.42	2
021	Cobalt (ppm)	0619	2.615	0.0500	4.671	0.8845	0.2509	25	-2.32	021.41	0
021	Cobalt (ppm)	0043	3.400	0.8200	4.671	0.8845	0.2509	25	-1.44	021.43	0
021	Cobalt (ppm)	0171	3.446	0.1210	4.671	0.8845	0.2509	25	-1.39	021.41	0
021	Cobalt (ppm)	0199	3.530	0.2400	4.671	0.8845	0.2509	25	-1.29	021.53	0
021	Cobalt (ppm)	0693	3.654	0.5693	4.671	0.8845	0.2509	25	-1.15	021.42	0
021	Cobalt (ppm)	0345	4.045	0.1100	4.671	0.8845	0.2509	25	-0.71	021.43	0
021	Cobalt (ppm)	0066	4.073	0.1130	4.671	0.8845	0.2509	25	-0.68	021.43	0
021	Cobalt (ppm)	0939	4.410	0.0600	4.671	0.8845	0.2509	25	-0.30	021.53	0
021	Cobalt (ppm)	0682	4.500	0.0000	4.671	0.8845	0.2509	25	-0.19	021.43	0
021	Cobalt (ppm)	0407	4.610	0.0359	4.671	0.8845	0.2509	25	-0.07	021.41	0
021	Cobalt (ppm)	0169	4.660	0.1200	4.671	0.8845	0.2509	25	-0.01	021.43	0
021	Cobalt (ppm)	0918	4.686	0.1960	4.671	0.8845	0.2509	25	0.02	021.53	0
021	Cobalt (ppm)	0875	4.750	0.1000	4.671	0.8845	0.2509	25	0.09	021.43	0
021	Cobalt (ppm)	0038	4.754	0.0140	4.671	0.8845	0.2509	25	0.09	021.43	0
021	Cobalt (ppm)	0964	4.843	0.1569	4.671	0.8845	0.2509	25	0.19	021.43	0
021	Cobalt (ppm)	0148	4.945	0.3700	4.671	0.8845	0.2509	25	0.31	021.41	0
021	Cobalt (ppm)	0560	5.110	0.5400	4.671	0.8845	0.2509	25	0.50	021.52	0
021	Cobalt (ppm)	0553	5.110	0.0200	4.671	0.8845	0.2509	25	0.50	021.53	0
021	Cobalt (ppm)	0910	5.345	0.4700	4.671	0.8845	0.2509	25	0.76	021.52	0
021	Cobalt (ppm)	0164	5.350	0.1000	4.671	0.8845	0.2509	25	0.77	021.31	0
021	Cobalt (ppm)	0510	5.380	0.3400	4.671	0.8845	0.2509	25	0.80	021.43	0
021	Cobalt (ppm)	0407	5.422	0.4866	4.671	0.8845	0.2509	25	0.85	021.43	0
021	Cobalt (ppm)	0939	5.455	0.3300	4.671	0.8845	0.2509	25	0.89	021.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
021	Cobalt (ppm)	0353	5.965	0.2595	4.671	0.8845	0.2509	25	1.46	021.43	0
021	Cobalt (ppm)	0047	10.96	0.6500	4.671	0.8845	0.2509	25	7.10	021.52	0
022	Copper (ppm)	2129	12.55	0.8300	121.7	27.16	4.936	91	-4.02	022.42	0
022	Copper (ppm)	0620	31.39	0.1213	121.7	27.16	4.936	91	-3.33	022.31	0
022	Copper (ppm)	0004	41.88	1.980	121.7	27.16	4.936	91	-2.94	022.41	0
022	Copper (ppm)	2012	54.28	0.1800	121.7	27.16	4.936	91	-2.48	022.41	0
022	Copper (ppm)	0536	54.88	0.2500	121.7	27.16	4.936	91	-2.46	022.31	0
022	Copper (ppm)	0590	58.10	2.200	121.7	27.16	4.936	91	-2.34	022.99	0
022	Copper (ppm)	0622	67.50	1.000	121.7	27.16	4.936	91	-2.00	022.31	0
022	Copper (ppm)	0619	71.45	0.9000	121.7	27.16	4.936	91	-1.85	022.41	0
022	Copper (ppm)	0171	72.87	0.4000	121.7	27.16	4.936	91	-1.80	022.41	0
022	Copper (ppm)	0049	76.39	14.22	121.7	27.16	4.936	91	-1.67	022.41	0
022	Copper (ppm)	0019	81.14	10.76	121.7	27.16	4.936	91	-1.49	022.41	0
022	Copper (ppm)	0001	83.23	0.7800	121.7	27.16	4.936	91	-1.42	022.31	0
022	Copper (ppm)	0354	83.27	2.210	121.7	27.16	4.936	91	-1.42	022.41	0
022	Copper (ppm)	0948	83.48	0.9550	121.7	27.16	4.936	91	-1.41	022.33	0
022	Copper (ppm)	0164	84.00	10.00	121.7	27.16	4.936	91	-1.39	022.44	0
022	Copper (ppm)	0653	84.50	4.490	121.7	27.16	4.936	91	-1.37	022.41	0
022	Copper (ppm)	0520	86.00	2.000	121.7	27.16	4.936	91	-1.32	022.41	0
022	Copper (ppm)	0598	94.31	0.5900	121.7	27.16	4.936	91	-1.01	022.41	0
022	Copper (ppm)	0003	94.50	9.000	121.7	27.16	4.936	91	-1.00	022.41	0
022	Copper (ppm)	0229	96.12	4.990	121.7	27.16	4.936	91	-0.94	022.41	0
022	Copper (ppm)	0868	100.6	8.840	121.7	27.16	4.936	91	-0.78	022.31	0
022	Copper (ppm)	2196	101.2	0.0000	121.7	27.16	4.936	91	-0.76	022.31	0
022	Copper (ppm)	0074	102.0	8.000	121.7	27.16	4.936	91	-0.73	022.41	0
022	Copper (ppm)	0148	105.1	2.840	121.7	27.16	4.936	91	-0.61	022.41	0
022	Copper (ppm)	0358	105.1	5.610	121.7	27.16	4.936	91	-0.61	022.41	0
022	Copper (ppm)	0199	105.3	8.640	121.7	27.16	4.936	91	-0.61	022.53	0
022	Copper (ppm)	0098	105.7	5.400	121.7	27.16	4.936	91	-0.59	022.41	0
022	Copper (ppm)	0878	110.0	6.000	121.7	27.16	4.936	91	-0.43	022.41	0
022	Copper (ppm)	0563	111.3	2.596	121.7	27.16	4.936	91	-0.38	022.31	0
022	Copper (ppm)	0226	112.4	6.760	121.7	27.16	4.936	91	-0.34	022.41	0
022	Copper (ppm)	0038	113.9	6.180	121.7	27.16	4.936	91	-0.29	022.32	0
022	Copper (ppm)	0294	116.3	1.140	121.7	27.16	4.936	91	-0.20	022.42	0
022	Copper (ppm)	0169	120.5	5.000	121.7	27.16	4.936	91	-0.05	022.43	0
022	Copper (ppm)	0175	122.0	0.0000	121.7	27.16	4.936	91	0.01	022.31	0
022	Copper (ppm)	0939	122.5	1.000	121.7	27.16	4.936	91	0.03	022.31	0
022	Copper (ppm)	2246	122.6	1.215	121.7	27.16	4.936	91	0.03	022.33	0
022	Copper (ppm)	0910	124.5	7.000	121.7	27.16	4.936	91	0.10	022.41	0
022	Copper (ppm)	0884	124.9	1.500	121.7	27.16	4.936	91	0.12	022.31	0
022	Copper (ppm)	0918	125.5	0.0100	121.7	27.16	4.936	91	0.14	022.43	0
022	Copper (ppm)	0553	125.5	11.00	121.7	27.16	4.936	91	0.14	022.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
022	Copper (ppm)	2089	126.3	10.32	121.7	27.16	4.936	91	0.17	022.43	0
022	Copper (ppm)	0939	127.0	4.000	121.7	27.16	4.936	91	0.19	022.53	0
022	Copper (ppm)	0572	128.5	1.000	121.7	27.16	4.936	91	0.25	022.53	0
022	Copper (ppm)	0345	129.0	2.300	121.7	27.16	4.936	91	0.27	022.43	0
022	Copper (ppm)	0407	129.1	22.24	121.7	27.16	4.936	91	0.27	022.41	0
022	Copper (ppm)	0870	130.8	5.700	121.7	27.16	4.936	91	0.33	022.42	0
022	Copper (ppm)	0242	131.0	4.000	121.7	27.16	4.936	91	0.34	022.99	0
022	Copper (ppm)	0098	131.3	7.200	121.7	27.16	4.936	91	0.35	022.43	0
022	Copper (ppm)	0278	131.6	6.500	121.7	27.16	4.936	91	0.36	022.42	0
022	Copper (ppm)	0047	131.6	13.80	121.7	27.16	4.936	91	0.36	022.52	0
022	Copper (ppm)	0083	132.0	4.000	121.7	27.16	4.936	91	0.38	022.43	0
022	Copper (ppm)	0726	132.3	0.2000	121.7	27.16	4.936	91	0.39	022.42	0
022	Copper (ppm)	0848	132.5	1.900	121.7	27.16	4.936	91	0.40	022.41	0
022	Copper (ppm)	0263	132.9	0.7700	121.7	27.16	4.936	91	0.41	022.42	0
022	Copper (ppm)	0875	133.0	4.000	121.7	27.16	4.936	91	0.42	022.43	0
022	Copper (ppm)	0066	133.1	3.840	121.7	27.16	4.936	91	0.42	022.43	0
022	Copper (ppm)	0043	135.5	9.500	121.7	27.16	4.936	91	0.51	022.43	0
022	Copper (ppm)	0265	135.5	3.000	121.7	27.16	4.936	91	0.51	022.42	0
022	Copper (ppm)	0366	135.5	9.000	121.7	27.16	4.936	91	0.51	022.42	0
022	Copper (ppm)	0297	136.0	10.00	121.7	27.16	4.936	91	0.53	022.43	0
022	Copper (ppm)	0968	136.0	0.0000	121.7	27.16	4.936	91	0.53	022.43	0
022	Copper (ppm)	0505	136.4	0.2000	121.7	27.16	4.936	91	0.54	022.43	0
022	Copper (ppm)	0964	136.5	4.068	121.7	27.16	4.936	91	0.55	022.43	0
022	Copper (ppm)	0560	137.2	3.700	121.7	27.16	4.936	91	0.57	022.52	0
022	Copper (ppm)	0357	139.0	0.8300	121.7	27.16	4.936	91	0.64	022.42	0
022	Copper (ppm)	0202	139.3	19.23	121.7	27.16	4.936	91	0.65	022.43	0
022	Copper (ppm)	0017	139.4	7.775	121.7	27.16	4.936	91	0.65	022.43	0
022	Copper (ppm)	0100	139.5	5.000	121.7	27.16	4.936	91	0.65	022.99	0
022	Copper (ppm)	0190	139.7	1.400	121.7	27.16	4.936	91	0.66	022.42	0
022	Copper (ppm)	0512	140.6	3.600	121.7	27.16	4.936	91	0.69	022.43	0
022	Copper (ppm)	0035	141.0	2.000	121.7	27.16	4.936	91	0.71	022.43	0
022	Copper (ppm)	0693	141.0	6.205	121.7	27.16	4.936	91	0.71	022.42	0
022	Copper (ppm)	0027	141.1	1.161	121.7	27.16	4.936	91	0.71	022.43	0
022	Copper (ppm)	0010	142.2	7.500	121.7	27.16	4.936	91	0.75	022.33	0
022	Copper (ppm)	0407	142.3	8.335	121.7	27.16	4.936	91	0.76	022.43	0
022	Copper (ppm)	0914	142.8	21.24	121.7	27.16	4.936	91	0.78	022.53	0
022	Copper (ppm)	0610	143.5	7.000	121.7	27.16	4.936	91	0.80	022.43	0
022	Copper (ppm)	0009	144.0	18.00	121.7	27.16	4.936	91	0.82	022.42	0
022	Copper (ppm)	0510	144.0	0.0000	121.7	27.16	4.936	91	0.82	022.43	0
022	Copper (ppm)	0032	145.0	5.000	121.7	27.16	4.936	91	0.86	022.43	0
022	Copper (ppm)	0685	145.5	0.0100	121.7	27.16	4.936	91	0.88	022.42	0
022	Copper (ppm)	0035	146.0	4.000	121.7	27.16	4.936	91	0.89	022.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
022	Copper (ppm)	0682	148.8	0.0000	121.7	27.16	4.936	91	1.00	022.43	0
022	Copper (ppm)	0413	149.5	7.000	121.7	27.16	4.936	91	1.02	022.42	0
022	Copper (ppm)	2192	150.8	4.780	121.7	27.16	4.936	91	1.07	022.43	0
022	Copper (ppm)	0038	151.8	4.900	121.7	27.16	4.936	91	1.11	022.43	0
022	Copper (ppm)	0692	152.0	8.000	121.7	27.16	4.936	91	1.11	022.42	0
022	Copper (ppm)	2053	152.5	3.000	121.7	27.16	4.936	91	1.13	022.42	0
022	Copper (ppm)	0037	153.2	4.300	121.7	27.16	4.936	91	1.16	022.42	0
022	Copper (ppm)	0675	166.4	3.050	121.7	27.16	4.936	91	1.64	022.43	0
022	Copper (ppm)	0353	170.0	8.000	121.7	27.16	4.936	91	1.78	022.43	0
022	Copper (ppm)	0529	115.7	26.80	121.7	27.16	4.936	91	-0.22	022.31	1
022	Copper (ppm)	0042	147.5	29.00	121.7	27.16	4.936	91	0.95	022.42	1
024	Iodine (ppm)	0910	4.915	0.6990				3		024.52	0
024	Iodine (ppm)	0668	5.100	0.4800				3		024.53	0
024	Iodine (ppm)	0160	5.515	0.1900				3		024.99	0
025	Iron (ppm)	0278	166.5	41.00	657.8	100.1	20.58	83	-4.91	025.42	0
025	Iron (ppm)	0560	189.1	9.200	657.8	100.1	20.58	83	-4.68	025.52	0
025	Iron (ppm)	2129	191.7	2.000	657.8	100.1	20.58	83	-4.66	025.42	0
025	Iron (ppm)	0726	246.5	0.0000	657.8	100.1	20.58	83	-4.11	025.42	0
025	Iron (ppm)	0265	319.0	20.00	657.8	100.1	20.58	83	-3.38	025.42	0
025	Iron (ppm)	2053	321.0	22.00	657.8	100.1	20.58	83	-3.36	025.42	0
025	Iron (ppm)	0199	408.5	7.000	657.8	100.1	20.58	83	-2.49	025.53	0
025	Iron (ppm)	0968	410.5	1.000	657.8	100.1	20.58	83	-2.47	025.43	0
025	Iron (ppm)	0921	448.6	7.320	657.8	100.1	20.58	83	-2.09	025.31	0
025	Iron (ppm)	0190	467.4	8.070	657.8	100.1	20.58	83	-1.90	025.42	0
025	Iron (ppm)	0047	493.9	20.90	657.8	100.1	20.58	83	-1.64	025.52	0
025	Iron (ppm)	0619	513.5	7.000	657.8	100.1	20.58	83	-1.44	025.41	0
025	Iron (ppm)	0598	532.5	30.00	657.8	100.1	20.58	83	-1.25	025.41	0
025	Iron (ppm)	0294	543.4	4.900	657.8	100.1	20.58	83	-1.14	025.42	0
025	Iron (ppm)	0226	546.3	5.090	657.8	100.1	20.58	83	-1.11	025.41	0
025	Iron (ppm)	0620	551.2	47.56	657.8	100.1	20.58	83	-1.06	025.31	0
025	Iron (ppm)	0037	567.7	63.60	657.8	100.1	20.58	83	-0.90	025.42	0
025	Iron (ppm)	0169	571.5	33.00	657.8	100.1	20.58	83	-0.86	025.43	0
025	Iron (ppm)	0425	575.3	99.80	657.8	100.1	20.58	83	-0.82	025.43	0
025	Iron (ppm)	2012	576.6	17.19	657.8	100.1	20.58	83	-0.81	025.41	0
025	Iron (ppm)	0043	581.0	38.00	657.8	100.1	20.58	83	-0.77	025.43	0
025	Iron (ppm)	0042	601.5	9.000	657.8	100.1	20.58	83	-0.56	025.42	0
025	Iron (ppm)	0529	609.0	17.00	657.8	100.1	20.58	83	-0.49	025.31	0
025	Iron (ppm)	0003	610.5	59.00	657.8	100.1	20.58	83	-0.47	025.41	0
025	Iron (ppm)	0653	617.0	8.800	657.8	100.1	20.58	83	-0.41	025.41	0
025	Iron (ppm)	0918	622.4	12.47	657.8	100.1	20.58	83	-0.35	025.43	0
025	Iron (ppm)	0910	625.0	70.00	657.8	100.1	20.58	83	-0.33	025.41	0
025	Iron (ppm)	0354	637.2	10.40	657.8	100.1	20.58	83	-0.21	025.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
025	Iron (ppm)	0083	642.5	1.000	657.8	100.1	20.58	83	-0.15	025.43	0
025	Iron (ppm)	0098	645.0	4.100	657.8	100.1	20.58	83	-0.13	025.41	0
025	Iron (ppm)	0505	647.3	52.00	657.8	100.1	20.58	83	-0.10	025.43	0
025	Iron (ppm)	0693	648.3	5.762	657.8	100.1	20.58	83	-0.09	025.42	0
025	Iron (ppm)	0345	648.5	7.000	657.8	100.1	20.58	83	-0.09	025.43	0
025	Iron (ppm)	0004	652.5	5.120	657.8	100.1	20.58	83	-0.05	025.41	0
025	Iron (ppm)	0032	654.9	29.80	657.8	100.1	20.58	83	-0.03	025.43	0
025	Iron (ppm)	0675	655.6	12.91	657.8	100.1	20.58	83	-0.02	025.43	0
025	Iron (ppm)	0878	658.5	9.000	657.8	100.1	20.58	83	0.01	025.41	0
025	Iron (ppm)	0413	672.0	46.00	657.8	100.1	20.58	83	0.14	025.42	0
025	Iron (ppm)	0001	674.8	11.10	657.8	100.1	20.58	83	0.17	025.31	0
025	Iron (ppm)	0366	675.0	52.00	657.8	100.1	20.58	83	0.17	025.42	0
025	Iron (ppm)	0358	678.3	19.44	657.8	100.1	20.58	83	0.20	025.41	0
025	Iron (ppm)	2246	679.3	2.242	657.8	100.1	20.58	83	0.22	025.33	0
025	Iron (ppm)	0610	680.5	41.00	657.8	100.1	20.58	83	0.23	025.43	0
025	Iron (ppm)	0066	680.6	22.27	657.8	100.1	20.58	83	0.23	025.43	0
025	Iron (ppm)	0553	684.0	6.000	657.8	100.1	20.58	83	0.26	025.53	0
025	Iron (ppm)	0263	689.7	4.610	657.8	100.1	20.58	83	0.32	025.42	0
025	Iron (ppm)	0171	692.5	1.800	657.8	100.1	20.58	83	0.35	025.41	0
025	Iron (ppm)	0870	693.9	13.10	657.8	100.1	20.58	83	0.36	025.42	0
025	Iron (ppm)	0407	695.1	71.31	657.8	100.1	20.58	83	0.37	025.41	0
025	Iron (ppm)	0164	697.5	31.00	657.8	100.1	20.58	83	0.40	025.41	0
025	Iron (ppm)	0074	698.0	10.00	657.8	100.1	20.58	83	0.40	025.41	0
025	Iron (ppm)	0563	699.0	7.711	657.8	100.1	20.58	83	0.41	025.31	0
025	Iron (ppm)	0868	699.5	7.000	657.8	100.1	20.58	83	0.42	025.31	0
025	Iron (ppm)	0692	700.0	48.00	657.8	100.1	20.58	83	0.42	025.42	0
025	Iron (ppm)	0297	701.5	23.00	657.8	100.1	20.58	83	0.44	025.43	0
025	Iron (ppm)	0229	702.8	11.60	657.8	100.1	20.58	83	0.45	025.41	0
025	Iron (ppm)	0098	703.4	22.80	657.8	100.1	20.58	83	0.46	025.43	0
025	Iron (ppm)	0939	704.5	5.000	657.8	100.1	20.58	83	0.47	025.31	0
025	Iron (ppm)	2196	705.4	0.0000	657.8	100.1	20.58	83	0.48	025.31	0
025	Iron (ppm)	0242	706.0	6.000	657.8	100.1	20.58	83	0.48	025.99	0
025	Iron (ppm)	0017	710.2	20.07	657.8	100.1	20.58	83	0.52	025.43	0
025	Iron (ppm)	0049	711.5	1.200	657.8	100.1	20.58	83	0.54	025.41	0
025	Iron (ppm)	0964	718.2	1.759	657.8	100.1	20.58	83	0.60	025.43	0
025	Iron (ppm)	0014	719.0	6.000	657.8	100.1	20.58	83	0.61	025.31	0
025	Iron (ppm)	0019	720.5	23.60	657.8	100.1	20.58	83	0.63	025.41	0
025	Iron (ppm)	0353	723.0	10.00	657.8	100.1	20.58	83	0.65	025.43	0
025	Iron (ppm)	0027	726.5	5.329	657.8	100.1	20.58	83	0.69	025.43	0
025	Iron (ppm)	0682	737.8	0.0000	657.8	100.1	20.58	83	0.80	025.43	0
025	Iron (ppm)	0622	743.4	37.59	657.8	100.1	20.58	83	0.85	025.31	0
025	Iron (ppm)	0884	748.9	2.200	657.8	100.1	20.58	83	0.91	025.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
025	Iron (ppm)	0175	749.0	14.00	657.8	100.1	20.58	83	0.91	025.31	0
025	Iron (ppm)	0510	751.0	2.000	657.8	100.1	20.58	83	0.93	025.43	0
025	Iron (ppm)	0512	757.2	78.50	657.8	100.1	20.58	83	0.99	025.43	0
025	Iron (ppm)	0148	758.2	6.940	657.8	100.1	20.58	83	1.00	025.41	0
025	Iron (ppm)	0536	761.7	35.10	657.8	100.1	20.58	83	1.04	025.31	0
025	Iron (ppm)	0100	762.5	3.000	657.8	100.1	20.58	83	1.05	025.99	0
025	Iron (ppm)	0407	769.4	51.77	657.8	100.1	20.58	83	1.12	025.43	0
025	Iron (ppm)	0939	774.5	11.00	657.8	100.1	20.58	83	1.17	025.53	0
025	Iron (ppm)	2192	795.0	45.73	657.8	100.1	20.58	83	1.37	025.43	0
025	Iron (ppm)	0202	802.8	31.09	657.8	100.1	20.58	83	1.45	025.43	0
025	Iron (ppm)	0038	809.5	15.00	657.8	100.1	20.58	83	1.52	025.43	0
025	Iron (ppm)	0520	820.0	6.000	657.8	100.1	20.58	83	1.62	025.41	0
025	Iron (ppm)	0875	1,054	39.00	657.8	100.1	20.58	83	3.95	025.43	0
025	Iron (ppm)	0123	503.0	108.0	657.8	100.1	20.58	83	-1.55	025.41	1
025	Iron (ppm)	2089	560.3	129.5	657.8	100.1	20.58	83	-0.97	025.43	1
027	Magnesium (%)	0889	0.3175	0.0050	0.3746	0.0172	0.0099	99	-3.32	027.99	0
027	Magnesium (%)	0505	0.3300	0.0200	0.3746	0.0172	0.0099	99	-2.59	027.43	0
027	Magnesium (%)	0353	0.3350	0.0100	0.3746	0.0172	0.0099	99	-2.30	027.43	0
027	Magnesium (%)	0875	0.3390	0.0060	0.3746	0.0172	0.0099	99	-2.07	027.43	0
027	Magnesium (%)	0004	0.3400	0.0200	0.3746	0.0172	0.0099	99	-2.01	027.41	0
027	Magnesium (%)	0089	0.3400	0.0000	0.3746	0.0172	0.0099	99	-2.01	027.43	0
027	Magnesium (%)	0038	0.3408	0.0100	0.3746	0.0172	0.0099	99	-1.97	027.31	0
027	Magnesium (%)	0884	0.3450	0.0100	0.3746	0.0172	0.0099	99	-1.72	027.31	0
027	Magnesium (%)	2012	0.3500	0.0000	0.3746	0.0172	0.0099	99	-1.43	027.41	0
027	Magnesium (%)	0357	0.3525	0.0108	0.3746	0.0172	0.0099	99	-1.29	027.42	0
027	Magnesium (%)	0529	0.3540	0.0000	0.3746	0.0172	0.0099	99	-1.20	027.31	0
027	Magnesium (%)	0229	0.3540	0.0100	0.3746	0.0172	0.0099	99	-1.20	027.41	0
027	Magnesium (%)	0407	0.3549	0.0125	0.3746	0.0172	0.0099	99	-1.15	027.43	0
027	Magnesium (%)	0049	0.3550	0.0100	0.3746	0.0172	0.0099	99	-1.14	027.41	0
027	Magnesium (%)	0042	0.3550	0.0400	0.3746	0.0172	0.0099	99	-1.14	027.42	0
027	Magnesium (%)	0294	0.3550	0.0100	0.3746	0.0172	0.0099	99	-1.14	027.42	0
027	Magnesium (%)	0083	0.3550	0.0100	0.3746	0.0172	0.0099	99	-1.14	027.43	0
027	Magnesium (%)	0202	0.3560	0.0140	0.3746	0.0172	0.0099	99	-1.08	027.43	0
027	Magnesium (%)	0693	0.3596	0.0033	0.3746	0.0172	0.0099	99	-0.88	027.42	0
027	Magnesium (%)	0536	0.3600	0.0000	0.3746	0.0172	0.0099	99	-0.85	027.31	0
027	Magnesium (%)	0169	0.3600	0.0200	0.3746	0.0172	0.0099	99	-0.85	027.32	0
027	Magnesium (%)	0043	0.3605	0.0010	0.3746	0.0172	0.0099	99	-0.82	027.43	0
027	Magnesium (%)	0914	0.3610	0.0180	0.3746	0.0172	0.0099	99	-0.79	027.43	0
027	Magnesium (%)	0345	0.3640	0.0040	0.3746	0.0172	0.0099	99	-0.62	027.43	0
027	Magnesium (%)	0620	0.3643	0.0129	0.3746	0.0172	0.0099	99	-0.60	027.31	0
027	Magnesium (%)	0175	0.3650	0.0100	0.3746	0.0172	0.0099	99	-0.56	027.31	0
027	Magnesium (%)	0868	0.3650	0.0100	0.3746	0.0172	0.0099	99	-0.56	027.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
027	Magnesium (%)	0939	0.3650	0.0100	0.3746	0.0172	0.0099	99	-0.56	027.31	0
027	Magnesium (%)	0003	0.3650	0.0100	0.3746	0.0172	0.0099	99	-0.56	027.41	0
027	Magnesium (%)	0265	0.3650	0.0100	0.3746	0.0172	0.0099	99	-0.56	027.42	0
027	Magnesium (%)	0297	0.3650	0.0100	0.3746	0.0172	0.0099	99	-0.56	027.43	0
027	Magnesium (%)	0242	0.3650	0.0100	0.3746	0.0172	0.0099	99	-0.56	027.99	0
027	Magnesium (%)	0598	0.3659	0.0062	0.3746	0.0172	0.0099	99	-0.51	027.41	0
027	Magnesium (%)	0870	0.3659	0.0026	0.3746	0.0172	0.0099	99	-0.51	027.42	0
027	Magnesium (%)	0142	0.3660	0.0047	0.3746	0.0172	0.0099	99	-0.50	027.31	0
027	Magnesium (%)	0001	0.3676	0.0096	0.3746	0.0172	0.0099	99	-0.41	027.31	0
027	Magnesium (%)	0910	0.3700	0.0200	0.3746	0.0172	0.0099	99	-0.27	027.41	0
027	Magnesium (%)	0035	0.3700	0.0000	0.3746	0.0172	0.0099	99	-0.27	027.42	0
027	Magnesium (%)	0726	0.3700	0.0000	0.3746	0.0172	0.0099	99	-0.27	027.42	0
027	Magnesium (%)	0164	0.3715	0.0050	0.3746	0.0172	0.0099	99	-0.18	027.44	0
027	Magnesium (%)	2246	0.3718	0.0043	0.3746	0.0172	0.0099	99	-0.17	027.33	0
027	Magnesium (%)	0968	0.3720	0.0020	0.3746	0.0172	0.0099	99	-0.15	027.43	0
027	Magnesium (%)	2146	0.3750	0.0100	0.3746	0.0172	0.0099	99	0.02	027.31	0
027	Magnesium (%)	0019	0.3750	0.0100	0.3746	0.0172	0.0099	99	0.02	027.41	0
027	Magnesium (%)	0878	0.3750	0.0100	0.3746	0.0172	0.0099	99	0.02	027.41	0
027	Magnesium (%)	0278	0.3750	0.0100	0.3746	0.0172	0.0099	99	0.02	027.42	0
027	Magnesium (%)	0035	0.3750	0.0100	0.3746	0.0172	0.0099	99	0.02	027.43	0
027	Magnesium (%)	0510	0.3750	0.0100	0.3746	0.0172	0.0099	99	0.02	027.43	0
027	Magnesium (%)	0590	0.3750	0.0100	0.3746	0.0172	0.0099	99	0.02	027.99	0
027	Magnesium (%)	0622	0.3753	0.0021	0.3746	0.0172	0.0099	99	0.04	027.31	0
027	Magnesium (%)	0074	0.3755	0.0050	0.3746	0.0172	0.0099	99	0.05	027.41	0
027	Magnesium (%)	0918	0.3755	0.0090	0.3746	0.0172	0.0099	99	0.05	027.43	0
027	Magnesium (%)	0964	0.3762	0.0162	0.3746	0.0172	0.0099	99	0.09	027.43	0
027	Magnesium (%)	2196	0.3800	0.0000	0.3746	0.0172	0.0099	99	0.31	027.31	0
027	Magnesium (%)	0098	0.3800	0.0200	0.3746	0.0172	0.0099	99	0.31	027.41	0
027	Magnesium (%)	0226	0.3800	0.0000	0.3746	0.0172	0.0099	99	0.31	027.41	0
027	Magnesium (%)	0354	0.3800	0.0000	0.3746	0.0172	0.0099	99	0.31	027.41	0
027	Magnesium (%)	0413	0.3800	0.0200	0.3746	0.0172	0.0099	99	0.31	027.41	0
027	Magnesium (%)	0009	0.3800	0.0200	0.3746	0.0172	0.0099	99	0.31	027.42	0
027	Magnesium (%)	0190	0.3800	0.0200	0.3746	0.0172	0.0099	99	0.31	027.42	0
027	Magnesium (%)	0692	0.3800	0.0200	0.3746	0.0172	0.0099	99	0.31	027.42	0
027	Magnesium (%)	0098	0.3800	0.0000	0.3746	0.0172	0.0099	99	0.31	027.43	0
027	Magnesium (%)	2089	0.3800	0.0000	0.3746	0.0172	0.0099	99	0.31	027.43	0
027	Magnesium (%)	0199	0.3800	0.0000	0.3746	0.0172	0.0099	99	0.31	027.53	0
027	Magnesium (%)	0154	0.3814	0.0205	0.3746	0.0172	0.0099	99	0.39	027.52	0
027	Magnesium (%)	2144	0.3814	0.0032	0.3746	0.0172	0.0099	99	0.39	027.99	0
027	Magnesium (%)	0066	0.3820	0.0080	0.3746	0.0172	0.0099	99	0.43	027.43	0
027	Magnesium (%)	0263	0.3847	0.0030	0.3746	0.0172	0.0099	99	0.59	027.42	0
027	Magnesium (%)	0010	0.3850	0.0100	0.3746	0.0172	0.0099	99	0.60	027.33	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
027	Magnesium (%)	0148	0.3850	0.0100	0.3746	0.0172	0.0099	99	0.60	027.41	0
027	Magnesium (%)	0358	0.3850	0.0300	0.3746	0.0172	0.0099	99	0.60	027.41	0
027	Magnesium (%)	0610	0.3850	0.0100	0.3746	0.0172	0.0099	99	0.60	027.43	0
027	Magnesium (%)	0675	0.3850	0.0100	0.3746	0.0172	0.0099	99	0.60	027.43	0
027	Magnesium (%)	2192	0.3850	0.0100	0.3746	0.0172	0.0099	99	0.60	027.43	0
027	Magnesium (%)	0939	0.3850	0.0300	0.3746	0.0172	0.0099	99	0.60	027.53	0
027	Magnesium (%)	0560	0.3860	0.0408	0.3746	0.0172	0.0099	99	0.66	027.52	0
027	Magnesium (%)	0553	0.3860	0.0020	0.3746	0.0172	0.0099	99	0.66	027.53	0
027	Magnesium (%)	0948	0.3865	0.0050	0.3746	0.0172	0.0099	99	0.69	027.33	0
027	Magnesium (%)	0572	0.3870	0.0200	0.3746	0.0172	0.0099	99	0.72	027.53	0
027	Magnesium (%)	0653	0.3880	0.0020	0.3746	0.0172	0.0099	99	0.78	027.41	0
027	Magnesium (%)	0027	0.3880	0.0060	0.3746	0.0172	0.0099	99	0.78	027.43	0
027	Magnesium (%)	0006	0.3885	0.0050	0.3746	0.0172	0.0099	99	0.81	027.43	0
027	Magnesium (%)	0171	0.3900	0.0000	0.3746	0.0172	0.0099	99	0.89	027.41	0
027	Magnesium (%)	0425	0.3900	0.0200	0.3746	0.0172	0.0099	99	0.89	027.43	0
027	Magnesium (%)	0139	0.3903	0.0020	0.3746	0.0172	0.0099	99	0.91	027.41	0
027	Magnesium (%)	0512	0.3925	0.0193	0.3746	0.0172	0.0099	99	1.04	027.43	0
027	Magnesium (%)	0038	0.3925	0.0010	0.3746	0.0172	0.0099	99	1.04	027.43	0
027	Magnesium (%)	0619	0.3945	0.0050	0.3746	0.0172	0.0099	99	1.16	027.41	0
027	Magnesium (%)	0366	0.3950	0.0300	0.3746	0.0172	0.0099	99	1.19	027.42	0
027	Magnesium (%)	2146	0.3950	0.0100	0.3746	0.0172	0.0099	99	1.19	027.51	0
027	Magnesium (%)	0017	0.3952	0.0061	0.3746	0.0172	0.0099	99	1.19	027.43	0
027	Magnesium (%)	0563	0.4024	0.0177	0.3746	0.0172	0.0099	99	1.61	027.31	0
027	Magnesium (%)	0032	0.4037	0.0019	0.3746	0.0172	0.0099	99	1.69	027.43	0
027	Magnesium (%)	0100	0.4050	0.0100	0.3746	0.0172	0.0099	99	1.77	027.99	0
027	Magnesium (%)	0037	0.4090	0.0020	0.3746	0.0172	0.0099	99	2.00	027.42	0
027	Magnesium (%)	0047	0.4100	0.0200	0.3746	0.0172	0.0099	99	2.06	027.52	0
027	Magnesium (%)	0685	0.4250	0.0100	0.3746	0.0172	0.0099	99	2.93	027.42	0
027	Magnesium (%)	0682	0.4400	0.0000	0.3746	0.0172	0.0099	99	3.80	027.43	0
027	Magnesium (%)	0520	0.5440	0.0160	0.3746	0.0172	0.0099	99	9.85	027.41	0
027	Magnesium (%)	0407	0.3515	0.0528	0.3746	0.0172	0.0099	99	-1.34	027.41	1
027	Magnesium (%)	0596	0.3650	0.0500	0.3746	0.0172	0.0099	99	-0.56	027.31	1
027	Magnesium (%)	2129	0.1153	0.0014	0.3746	0.0172	0.0099	99	-15.07	027.42	2
028	Manganese (ppm)	0226	328.6	8.440	406.9	33.57	10.12	88	-2.33	028.41	0
028	Manganese (ppm)	2012	336.3	9.850	406.9	33.57	10.12	88	-2.10	028.41	0
028	Manganese (ppm)	0619	341.0	4.000	406.9	33.57	10.12	88	-1.96	028.41	0
028	Manganese (ppm)	0653	349.2	2.900	406.9	33.57	10.12	88	-1.72	028.41	0
028	Manganese (ppm)	0910	350.5	1.000	406.9	33.57	10.12	88	-1.68	028.41	0
028	Manganese (ppm)	0199	351.3	11.30	406.9	33.57	10.12	88	-1.66	028.53	0
028	Manganese (ppm)	0049	356.8	28.90	406.9	33.57	10.12	88	-1.49	028.41	0
028	Manganese (ppm)	0004	365.1	1.290	406.9	33.57	10.12	88	-1.25	028.41	0
028	Manganese (ppm)	2196	367.2	0.0000	406.9	33.57	10.12	88	-1.18	028.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
028	Manganese (ppm)	0169	367.5	13.00	406.9	33.57	10.12	88	-1.17	028.43	0
028	Manganese (ppm)	0354	369.6	0.9900	406.9	33.57	10.12	88	-1.11	028.41	0
028	Manganese (ppm)	0685	370.5	0.0100	406.9	33.57	10.12	88	-1.08	028.42	0
028	Manganese (ppm)	0294	371.2	1.320	406.9	33.57	10.12	88	-1.06	028.42	0
028	Manganese (ppm)	0164	371.5	1.000	406.9	33.57	10.12	88	-1.06	028.44	0
028	Manganese (ppm)	0914	374.0	8.000	406.9	33.57	10.12	88	-0.98	028.43	0
028	Manganese (ppm)	0590	375.0	6.000	406.9	33.57	10.12	88	-0.95	028.99	0
028	Manganese (ppm)	0536	377.1	5.900	406.9	33.57	10.12	88	-0.89	028.31	0
028	Manganese (ppm)	0003	378.5	11.00	406.9	33.57	10.12	88	-0.85	028.41	0
028	Manganese (ppm)	0948	378.9	3.630	406.9	33.57	10.12	88	-0.84	028.33	0
028	Manganese (ppm)	0407	379.8	45.91	406.9	33.57	10.12	88	-0.81	028.41	0
028	Manganese (ppm)	0875	380.5	1.000	406.9	33.57	10.12	88	-0.79	028.43	0
028	Manganese (ppm)	0066	380.6	13.27	406.9	33.57	10.12	88	-0.79	028.43	0
028	Manganese (ppm)	0098	383.0	18.40	406.9	33.57	10.12	88	-0.71	028.41	0
028	Manganese (ppm)	0074	383.5	9.000	406.9	33.57	10.12	88	-0.70	028.41	0
028	Manganese (ppm)	0620	384.8	4.120	406.9	33.57	10.12	88	-0.66	028.31	0
028	Manganese (ppm)	0229	385.9	18.50	406.9	33.57	10.12	88	-0.63	028.41	0
028	Manganese (ppm)	0001	390.1	16.10	406.9	33.57	10.12	88	-0.50	028.31	0
028	Manganese (ppm)	0693	392.3	8.860	406.9	33.57	10.12	88	-0.44	028.42	0
028	Manganese (ppm)	0505	393.1	1.100	406.9	33.57	10.12	88	-0.41	028.43	0
028	Manganese (ppm)	0553	394.0	8.000	406.9	33.57	10.12	88	-0.39	028.53	0
028	Manganese (ppm)	0878	398.5	9.000	406.9	33.57	10.12	88	-0.25	028.41	0
028	Manganese (ppm)	0884	398.8	3.700	406.9	33.57	10.12	88	-0.24	028.31	0
028	Manganese (ppm)	0425	399.4	2.000	406.9	33.57	10.12	88	-0.22	028.43	0
028	Manganese (ppm)	0572	399.5	9.000	406.9	33.57	10.12	88	-0.22	028.53	0
028	Manganese (ppm)	0019	400.2	14.40	406.9	33.57	10.12	88	-0.20	028.41	0
028	Manganese (ppm)	0175	401.0	18.00	406.9	33.57	10.12	88	-0.18	028.31	0
028	Manganese (ppm)	0263	401.4	5.030	406.9	33.57	10.12	88	-0.17	028.42	0
028	Manganese (ppm)	0939	401.5	9.000	406.9	33.57	10.12	88	-0.16	028.31	0
028	Manganese (ppm)	0939	401.5	3.000	406.9	33.57	10.12	88	-0.16	028.53	0
028	Manganese (ppm)	0968	403.0	0.0000	406.9	33.57	10.12	88	-0.12	028.43	0
028	Manganese (ppm)	0345	403.2	9.900	406.9	33.57	10.12	88	-0.11	028.43	0
028	Manganese (ppm)	0563	403.8	2.340	406.9	33.57	10.12	88	-0.09	028.31	0
028	Manganese (ppm)	0038	405.1	1.000	406.9	33.57	10.12	88	-0.05	028.32	0
028	Manganese (ppm)	0043	405.9	21.20	406.9	33.57	10.12	88	-0.03	028.43	0
028	Manganese (ppm)	0529	408.7	5.300	406.9	33.57	10.12	88	0.05	028.31	0
028	Manganese (ppm)	0598	408.7	19.40	406.9	33.57	10.12	88	0.05	028.41	0
028	Manganese (ppm)	0242	409.5	3.000	406.9	33.57	10.12	88	0.08	028.99	0
028	Manganese (ppm)	0083	410.5	5.000	406.9	33.57	10.12	88	0.11	028.43	0
028	Manganese (ppm)	0297	412.0	20.00	406.9	33.57	10.12	88	0.15	028.43	0
028	Manganese (ppm)	0610	413.0	8.000	406.9	33.57	10.12	88	0.18	028.43	0
028	Manganese (ppm)	0278	413.5	1.000	406.9	33.57	10.12	88	0.20	028.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
028	Manganese (ppm)	0918	414.2	5.210	406.9	33.57	10.12	88	0.22	028.43	0
028	Manganese (ppm)	0171	414.3	23.60	406.9	33.57	10.12	88	0.22	028.41	0
028	Manganese (ppm)	0868	414.5	11.00	406.9	33.57	10.12	88	0.23	028.31	0
028	Manganese (ppm)	0366	417.0	18.00	406.9	33.57	10.12	88	0.30	028.42	0
028	Manganese (ppm)	0413	417.0	24.00	406.9	33.57	10.12	88	0.30	028.42	0
028	Manganese (ppm)	0675	419.3	8.750	406.9	33.57	10.12	88	0.37	028.43	0
028	Manganese (ppm)	0035	421.0	10.00	406.9	33.57	10.12	88	0.42	028.42	0
028	Manganese (ppm)	0682	423.2	0.0000	406.9	33.57	10.12	88	0.48	028.43	0
028	Manganese (ppm)	0353	423.5	1.000	406.9	33.57	10.12	88	0.49	028.43	0
028	Manganese (ppm)	0512	424.0	14.40	406.9	33.57	10.12	88	0.51	028.43	0
028	Manganese (ppm)	0098	424.4	6.300	406.9	33.57	10.12	88	0.52	028.43	0
028	Manganese (ppm)	0035	424.5	11.00	406.9	33.57	10.12	88	0.52	028.43	0
028	Manganese (ppm)	0964	424.9	18.19	406.9	33.57	10.12	88	0.53	028.43	0
028	Manganese (ppm)	0358	425.0	11.25	406.9	33.57	10.12	88	0.54	028.41	0
028	Manganese (ppm)	0017	425.1	11.04	406.9	33.57	10.12	88	0.54	028.43	0
028	Manganese (ppm)	0357	426.5	2.002	406.9	33.57	10.12	88	0.58	028.42	0
028	Manganese (ppm)	0190	427.0	0.3300	406.9	33.57	10.12	88	0.60	028.42	0
028	Manganese (ppm)	0148	429.7	24.02	406.9	33.57	10.12	88	0.68	028.41	0
028	Manganese (ppm)	2089	431.0	20.59	406.9	33.57	10.12	88	0.72	028.43	0
028	Manganese (ppm)	0009	431.0	30.00	406.9	33.57	10.12	88	0.72	028.42	0
028	Manganese (ppm)	0870	432.3	5.400	406.9	33.57	10.12	88	0.76	028.42	0
028	Manganese (ppm)	0510	432.5	5.000	406.9	33.57	10.12	88	0.76	028.43	0
028	Manganese (ppm)	0692	435.0	32.00	406.9	33.57	10.12	88	0.84	028.42	0
028	Manganese (ppm)	0726	437.2	0.2000	406.9	33.57	10.12	88	0.90	028.42	0
028	Manganese (ppm)	0407	445.1	6.730	406.9	33.57	10.12	88	1.14	028.43	0
028	Manganese (ppm)	0038	448.0	2.000	406.9	33.57	10.12	88	1.22	028.43	0
028	Manganese (ppm)	0265	449.5	9.000	406.9	33.57	10.12	88	1.27	028.42	0
028	Manganese (ppm)	2053	452.0	8.000	406.9	33.57	10.12	88	1.34	028.42	0
028	Manganese (ppm)	0047	454.6	3.400	406.9	33.57	10.12	88	1.42	028.52	0
028	Manganese (ppm)	0027	457.4	4.052	406.9	33.57	10.12	88	1.50	028.43	0
028	Manganese (ppm)	2144	458.4	8.120	406.9	33.57	10.12	88	1.53	028.99	0
028	Manganese (ppm)	0037	461.1	1.700	406.9	33.57	10.12	88	1.61	028.42	0
028	Manganese (ppm)	0202	463.7	30.17	406.9	33.57	10.12	88	1.69	028.43	0
028	Manganese (ppm)	0520	472.5	5.000	406.9	33.57	10.12	88	1.95	028.41	0
028	Manganese (ppm)	0100	472.5	41.00	406.9	33.57	10.12	88	1.95	028.99	0
028	Manganese (ppm)	0596	482.0	11.07	406.9	33.57	10.12	88	2.24	028.31	0
028	Manganese (ppm)	0032	498.1	35.30	406.9	33.57	10.12	88	2.71	028.43	0
028	Manganese (ppm)	0042	398.5	81.00	406.9	33.57	10.12	88	-0.25	028.42	1
028	Manganese (ppm)	0560	453.3	62.80	406.9	33.57	10.12	88	1.38	028.52	1
028	Manganese (ppm)	2129	49.81	0.5000	406.9	33.57	10.12	88	-10.64	028.42	2
031	Phosphorus (%)	0889	0.7200	0.0100	1.160	0.0594	0.0334	128	-7.41	031.99	0
031	Phosphorus (%)	0354	0.7450	0.0100	1.160	0.0594	0.0334	128	-6.99	031.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	0154	0.9789	0.0046	1.160	0.0594	0.0334	128	-3.05	031.52	0
031	Phosphorus (%)	0852	0.9900	0.0200	1.160	0.0594	0.0334	128	-2.87	031.99	0
031	Phosphorus (%)	2259	1.021	0.0030	1.160	0.0594	0.0334	128	-2.35	031.41	0
031	Phosphorus (%)	0047	1.025	0.1500	1.160	0.0594	0.0334	128	-2.28	031.42	0
031	Phosphorus (%)	0536	1.040	0.0400	1.160	0.0594	0.0334	128	-2.02	031.06	0
031	Phosphorus (%)	0009	1.040	0.0000	1.160	0.0594	0.0334	128	-2.02	031.42	0
031	Phosphorus (%)	0018	1.055	0.0300	1.160	0.0594	0.0334	128	-1.77	031.01	0
031	Phosphorus (%)	2129	1.057	0.0280	1.160	0.0594	0.0334	128	-1.74	031.42	0
031	Phosphorus (%)	2295	1.065	0.0100	1.160	0.0594	0.0334	128	-1.60	031.00	0
031	Phosphorus (%)	0685	1.065	0.0100	1.160	0.0594	0.0334	128	-1.60	031.41	0
031	Phosphorus (%)	0910	1.065	0.0300	1.160	0.0594	0.0334	128	-1.60	031.41	0
031	Phosphorus (%)	0278	1.080	0.0400	1.160	0.0594	0.0334	128	-1.35	031.42	0
031	Phosphorus (%)	0357	1.083	0.0031	1.160	0.0594	0.0334	128	-1.31	031.42	0
031	Phosphorus (%)	0004	1.090	0.0200	1.160	0.0594	0.0334	128	-1.18	031.41	0
031	Phosphorus (%)	0875	1.090	0.0200	1.160	0.0594	0.0334	128	-1.18	031.43	0
031	Phosphorus (%)	0939	1.090	0.0600	1.160	0.0594	0.0334	128	-1.18	031.53	0
031	Phosphorus (%)	0560	1.094	0.1560	1.160	0.0594	0.0334	128	-1.11	031.52	0
031	Phosphorus (%)	0693	1.096	0.0252	1.160	0.0594	0.0334	128	-1.08	031.42	0
031	Phosphorus (%)	0199	1.097	0.0050	1.160	0.0594	0.0334	128	-1.07	031.53	0
031	Phosphorus (%)	0674	1.100	0.0000	1.160	0.0594	0.0334	128	-1.01	031.01	0
031	Phosphorus (%)	0676	1.100	0.0400	1.160	0.0594	0.0334	128	-1.01	031.99	0
031	Phosphorus (%)	0265	1.105	0.0300	1.160	0.0594	0.0334	128	-0.93	031.42	0
031	Phosphorus (%)	0001	1.105	0.0500	1.160	0.0594	0.0334	128	-0.93	031.43	0
031	Phosphorus (%)	0083	1.105	0.0100	1.160	0.0594	0.0334	128	-0.93	031.43	0
031	Phosphorus (%)	0505	1.105	0.0700	1.160	0.0594	0.0334	128	-0.93	031.43	0
031	Phosphorus (%)	2200	1.110	0.0800	1.160	0.0594	0.0334	128	-0.85	031.01	0
031	Phosphorus (%)	0169	1.115	0.0500	1.160	0.0594	0.0334	128	-0.76	031.03	0
031	Phosphorus (%)	0298	1.120	0.0200	1.160	0.0594	0.0334	128	-0.68	031.41	0
031	Phosphorus (%)	0610	1.120	0.0000	1.160	0.0594	0.0334	128	-0.68	031.43	0
031	Phosphorus (%)	0358	1.125	0.0900	1.160	0.0594	0.0334	128	-0.59	031.41	0
031	Phosphorus (%)	0297	1.125	0.0900	1.160	0.0594	0.0334	128	-0.59	031.43	0
031	Phosphorus (%)	0918	1.129	0.0360	1.160	0.0594	0.0334	128	-0.53	031.43	0
031	Phosphorus (%)	0620	1.130	0.0098	1.160	0.0594	0.0334	128	-0.52	031.01	0
031	Phosphorus (%)	0019	1.130	0.0200	1.160	0.0594	0.0334	128	-0.51	031.41	0
031	Phosphorus (%)	0848	1.130	0.0200	1.160	0.0594	0.0334	128	-0.51	031.41	0
031	Phosphorus (%)	0914	1.130	0.0800	1.160	0.0594	0.0334	128	-0.51	031.43	0
031	Phosphorus (%)	0142	1.132	0.0227	1.160	0.0594	0.0334	128	-0.47	031.01	0
031	Phosphorus (%)	0590	1.135	0.0100	1.160	0.0594	0.0334	128	-0.43	031.01	0
031	Phosphorus (%)	0175	1.135	0.0300	1.160	0.0594	0.0334	128	-0.43	031.01	0
031	Phosphorus (%)	0098	1.138	0.0480	1.160	0.0594	0.0334	128	-0.37	031.41	0
031	Phosphorus (%)	0884	1.140	0.0200	1.160	0.0594	0.0334	128	-0.34	031.01	0
031	Phosphorus (%)	0897	1.140	0.0000	1.160	0.0594	0.0334	128	-0.34	031.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	0144	1.145	0.0300	1.160	0.0594	0.0334	128	-0.26	031.41	0
031	Phosphorus (%)	0190	1.145	0.0100	1.160	0.0594	0.0334	128	-0.26	031.42	0
031	Phosphorus (%)	0010	1.145	0.0100	1.160	0.0594	0.0334	128	-0.26	031.43	0
031	Phosphorus (%)	0229	1.146	0.1080	1.160	0.0594	0.0334	128	-0.24	031.41	0
031	Phosphorus (%)	0596	1.150	0.0400	1.160	0.0594	0.0334	128	-0.17	031.01	0
031	Phosphorus (%)	2192	1.150	0.0400	1.160	0.0594	0.0334	128	-0.17	031.01	0
031	Phosphorus (%)	0226	1.150	0.0000	1.160	0.0594	0.0334	128	-0.17	031.41	0
031	Phosphorus (%)	2012	1.150	0.0200	1.160	0.0594	0.0334	128	-0.17	031.41	0
031	Phosphorus (%)	0242	1.150	0.0600	1.160	0.0594	0.0334	128	-0.17	031.99	0
031	Phosphorus (%)	0968	1.152	0.0090	1.160	0.0594	0.0334	128	-0.15	031.43	0
031	Phosphorus (%)	0964	1.152	0.0590	1.160	0.0594	0.0334	128	-0.14	031.43	0
031	Phosphorus (%)	0653	1.153	0.0100	1.160	0.0594	0.0334	128	-0.12	031.41	0
031	Phosphorus (%)	2246	1.155	0.0022	1.160	0.0594	0.0334	128	-0.10	031.01	0
031	Phosphorus (%)	2146	1.155	0.0300	1.160	0.0594	0.0334	128	-0.09	031.01	0
031	Phosphorus (%)	0035	1.155	0.0100	1.160	0.0594	0.0334	128	-0.09	031.42	0
031	Phosphorus (%)	0563	1.156	0.0064	1.160	0.0594	0.0334	128	-0.08	031.01	0
031	Phosphorus (%)	2181	1.160	0.0000	1.160	0.0594	0.0334	128	0.00	031.01	0
031	Phosphorus (%)	0123	1.160	0.0800	1.160	0.0594	0.0334	128	0.00	031.41	0
031	Phosphorus (%)	0425	1.160	0.1000	1.160	0.0594	0.0334	128	0.00	031.43	0
031	Phosphorus (%)	0893	1.165	0.0300	1.160	0.0594	0.0334	128	0.08	031.00	0
031	Phosphorus (%)	2109	1.169	0.0460	1.160	0.0594	0.0334	128	0.15	031.01	0
031	Phosphorus (%)	0366	1.170	0.0600	1.160	0.0594	0.0334	128	0.16	031.42	0
031	Phosphorus (%)	0675	1.170	0.0200	1.160	0.0594	0.0334	128	0.16	031.43	0
031	Phosphorus (%)	0100	1.170	0.0000	1.160	0.0594	0.0334	128	0.16	031.99	0
031	Phosphorus (%)	0066	1.171	0.0730	1.160	0.0594	0.0334	128	0.17	031.43	0
031	Phosphorus (%)	0407	1.172	0.0498	1.160	0.0594	0.0334	128	0.19	031.41	0
031	Phosphorus (%)	0037	1.172	0.0180	1.160	0.0594	0.0334	128	0.20	031.42	0
031	Phosphorus (%)	0622	1.173	0.0012	1.160	0.0594	0.0334	128	0.22	031.01	0
031	Phosphorus (%)	0027	1.175	0.0770	1.160	0.0594	0.0334	128	0.24	031.43	0
031	Phosphorus (%)	0687	1.175	0.0500	1.160	0.0594	0.0334	128	0.25	031.01	0
031	Phosphorus (%)	0171	1.175	0.0100	1.160	0.0594	0.0334	128	0.25	031.41	0
031	Phosphorus (%)	0572	1.175	0.0300	1.160	0.0594	0.0334	128	0.25	031.53	0
031	Phosphorus (%)	2196	1.177	0.0000	1.160	0.0594	0.0334	128	0.28	031.01	0
031	Phosphorus (%)	0878	1.180	0.0200	1.160	0.0594	0.0334	128	0.33	031.01	0
031	Phosphorus (%)	0049	1.180	0.0600	1.160	0.0594	0.0334	128	0.33	031.41	0
031	Phosphorus (%)	0692	1.180	0.0400	1.160	0.0594	0.0334	128	0.33	031.42	0
031	Phosphorus (%)	0682	1.180	0.0000	1.160	0.0594	0.0334	128	0.33	031.43	0
031	Phosphorus (%)	0036	1.184	0.0075	1.160	0.0594	0.0334	128	0.40	031.03	0
031	Phosphorus (%)	2089	1.185	0.0100	1.160	0.0594	0.0334	128	0.42	031.43	0
031	Phosphorus (%)	2009	1.186	0.0034	1.160	0.0594	0.0334	128	0.43	031.01	0
031	Phosphorus (%)	0881	1.188	0.0080	1.160	0.0594	0.0334	128	0.47	031.01	0
031	Phosphorus (%)	0139	1.188	0.0300	1.160	0.0594	0.0334	128	0.47	031.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	0868	1.190	0.0200	1.160	0.0594	0.0334	128	0.50	031.01	0
031	Phosphorus (%)	0939	1.190	0.0400	1.160	0.0594	0.0334	128	0.50	031.01	0
031	Phosphorus (%)	0413	1.190	0.0000	1.160	0.0594	0.0334	128	0.50	031.41	0
031	Phosphorus (%)	0726	1.190	0.0000	1.160	0.0594	0.0334	128	0.50	031.42	0
031	Phosphorus (%)	0035	1.190	0.0400	1.160	0.0594	0.0334	128	0.50	031.43	0
031	Phosphorus (%)	0164	1.190	0.0400	1.160	0.0594	0.0334	128	0.50	031.44	0
031	Phosphorus (%)	0036	1.190	0.0335	1.160	0.0594	0.0334	128	0.51	031.43	0
031	Phosphorus (%)	0512	1.192	0.0450	1.160	0.0594	0.0334	128	0.53	031.43	0
031	Phosphorus (%)	0017	1.193	0.0132	1.160	0.0594	0.0334	128	0.55	031.43	0
031	Phosphorus (%)	0001	1.193	0.0020	1.160	0.0594	0.0334	128	0.55	031.03	0
031	Phosphorus (%)	2062	1.193	0.0003	1.160	0.0594	0.0334	128	0.55	031.01	0
031	Phosphorus (%)	2152	1.195	0.0100	1.160	0.0594	0.0334	128	0.59	031.01	0
031	Phosphorus (%)	0345	1.195	0.0300	1.160	0.0594	0.0334	128	0.59	031.43	0
031	Phosphorus (%)	2146	1.195	0.0500	1.160	0.0594	0.0334	128	0.59	031.51	0
031	Phosphorus (%)	0529	1.200	0.0200	1.160	0.0594	0.0334	128	0.67	031.01	0
031	Phosphorus (%)	0553	1.200	0.1200	1.160	0.0594	0.0334	128	0.67	031.53	0
031	Phosphorus (%)	0098	1.201	0.0160	1.160	0.0594	0.0334	128	0.69	031.43	0
031	Phosphorus (%)	0038	1.204	0.0005	1.160	0.0594	0.0334	128	0.73	031.01	0
031	Phosphorus (%)	0948	1.207	0.0350	1.160	0.0594	0.0334	128	0.78	031.01	0
031	Phosphorus (%)	2144	1.210	0.0000	1.160	0.0594	0.0334	128	0.84	031.01	0
031	Phosphorus (%)	0263	1.211	0.0028	1.160	0.0594	0.0334	128	0.85	031.42	0
031	Phosphorus (%)	0870	1.214	0.0290	1.160	0.0594	0.0334	128	0.90	031.42	0
031	Phosphorus (%)	0202	1.217	0.1270	1.160	0.0594	0.0334	128	0.95	031.43	0
031	Phosphorus (%)	0006	1.217	0.0440	1.160	0.0594	0.0334	128	0.96	031.43	0
031	Phosphorus (%)	2006	1.222	0.0110	1.160	0.0594	0.0334	128	1.03	031.01	0
031	Phosphorus (%)	0074	1.225	0.0300	1.160	0.0594	0.0334	128	1.09	031.41	0
031	Phosphorus (%)	0294	1.225	0.0300	1.160	0.0594	0.0334	128	1.09	031.42	0
031	Phosphorus (%)	0014	1.230	0.1000	1.160	0.0594	0.0334	128	1.17	031.42	0
031	Phosphorus (%)	0033	1.230	0.1400	1.160	0.0594	0.0334	128	1.17	031.43	0
031	Phosphorus (%)	0038	1.234	0.0210	1.160	0.0594	0.0334	128	1.23	031.43	0
031	Phosphorus (%)	2022	1.245	0.0100	1.160	0.0594	0.0334	128	1.43	031.01	0
031	Phosphorus (%)	0407	1.250	0.0651	1.160	0.0594	0.0334	128	1.51	031.43	0
031	Phosphorus (%)	0089	1.250	0.0000	1.160	0.0594	0.0334	128	1.51	031.43	0
031	Phosphorus (%)	0510	1.250	0.0200	1.160	0.0594	0.0334	128	1.51	031.43	0
031	Phosphorus (%)	0148	1.255	0.0300	1.160	0.0594	0.0334	128	1.60	031.41	0
031	Phosphorus (%)	2193	1.260	0.0000	1.160	0.0594	0.0334	128	1.68	031.01	0
031	Phosphorus (%)	0353	1.260	0.0200	1.160	0.0594	0.0334	128	1.68	031.43	0
031	Phosphorus (%)	0598	1.270	0.0440	1.160	0.0594	0.0334	128	1.85	031.41	0
031	Phosphorus (%)	0619	1.300	0.0200	1.160	0.0594	0.0334	128	2.35	031.41	0
031	Phosphorus (%)	0032	1.311	0.0360	1.160	0.0594	0.0334	128	2.54	031.43	0
031	Phosphorus (%)	0520	1.500	0.0400	1.160	0.0594	0.0334	128	5.72	031.41	0
031	Phosphorus (%)	0003	1.535	0.1700	1.160	0.0594	0.0334	128	6.31	031.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	2199	1.105	0.1900	1.160	0.0594	0.0334	128	-0.93	031.01	1
031	Phosphorus (%)	0042	1.107	0.2460	1.160	0.0594	0.0334	128	-0.90	031.42	1
032	Potassium (%)	0004	0.6600	0.0200	0.8500	0.0580	0.0214	96	-3.27	032.41	0
032	Potassium (%)	2129	0.6971	0.0049	0.8500	0.0580	0.0214	96	-2.63	032.42	0
032	Potassium (%)	0229	0.7370	0.0280	0.8500	0.0580	0.0214	96	-1.95	032.41	0
032	Potassium (%)	0536	0.7450	0.0100	0.8500	0.0580	0.0214	96	-1.81	032.31	0
032	Potassium (%)	2012	0.7450	0.0100	0.8500	0.0580	0.0214	96	-1.81	032.41	0
032	Potassium (%)	0868	0.7500	0.0200	0.8500	0.0580	0.0214	96	-1.72	032.31	0
032	Potassium (%)	0154	0.7500	0.0074	0.8500	0.0580	0.0214	96	-1.72	032.52	0
032	Potassium (%)	0875	0.7505	0.0150	0.8500	0.0580	0.0214	96	-1.71	032.43	0
032	Potassium (%)	0590	0.7700	0.0200	0.8500	0.0580	0.0214	96	-1.38	032.99	0
032	Potassium (%)	0019	0.7750	0.0100	0.8500	0.0580	0.0214	96	-1.29	032.41	0
032	Potassium (%)	0572	0.7790	0.0500	0.8500	0.0580	0.0214	96	-1.22	032.53	0
032	Potassium (%)	0123	0.7800	0.0200	0.8500	0.0580	0.0214	96	-1.21	032.41	0
032	Potassium (%)	0358	0.7800	0.0600	0.8500	0.0580	0.0214	96	-1.21	032.41	0
032	Potassium (%)	0265	0.7800	0.0200	0.8500	0.0580	0.0214	96	-1.21	032.42	0
032	Potassium (%)	0619	0.7920	0.0100	0.8500	0.0580	0.0214	96	-1.00	032.41	0
032	Potassium (%)	0529	0.7950	0.0100	0.8500	0.0580	0.0214	96	-0.95	032.31	0
032	Potassium (%)	0563	0.7960	0.0221	0.8500	0.0580	0.0214	96	-0.93	032.31	0
032	Potassium (%)	0175	0.8050	0.0300	0.8500	0.0580	0.0214	96	-0.78	032.31	0
032	Potassium (%)	0144	0.8050	0.0300	0.8500	0.0580	0.0214	96	-0.78	032.41	0
032	Potassium (%)	0505	0.8050	0.0100	0.8500	0.0580	0.0214	96	-0.78	032.43	0
032	Potassium (%)	0889	0.8050	0.0100	0.8500	0.0580	0.0214	96	-0.78	032.99	0
032	Potassium (%)	0049	0.8100	0.0200	0.8500	0.0580	0.0214	96	-0.69	032.41	0
032	Potassium (%)	0074	0.8100	0.0000	0.8500	0.0580	0.0214	96	-0.69	032.41	0
032	Potassium (%)	0294	0.8100	0.0000	0.8500	0.0580	0.0214	96	-0.69	032.42	0
032	Potassium (%)	2192	0.8100	0.0200	0.8500	0.0580	0.0214	96	-0.69	032.43	0
032	Potassium (%)	2146	0.8150	0.0100	0.8500	0.0580	0.0214	96	-0.60	032.31	0
032	Potassium (%)	0878	0.8150	0.0100	0.8500	0.0580	0.0214	96	-0.60	032.41	0
032	Potassium (%)	0001	0.8160	0.0020	0.8500	0.0580	0.0214	96	-0.59	032.99	0
032	Potassium (%)	0164	0.8180	0.0300	0.8500	0.0580	0.0214	96	-0.55	032.44	0
032	Potassium (%)	0939	0.8200	0.0200	0.8500	0.0580	0.0214	96	-0.52	032.31	0
032	Potassium (%)	0598	0.8277	0.0162	0.8500	0.0580	0.0214	96	-0.38	032.41	0
032	Potassium (%)	0242	0.8300	0.0000	0.8500	0.0580	0.0214	96	-0.34	032.99	0
032	Potassium (%)	0038	0.8305	0.0400	0.8500	0.0580	0.0214	96	-0.34	032.31	0
032	Potassium (%)	0098	0.8350	0.0700	0.8500	0.0580	0.0214	96	-0.26	032.41	0
032	Potassium (%)	0083	0.8350	0.0100	0.8500	0.0580	0.0214	96	-0.26	032.43	0
032	Potassium (%)	0098	0.8350	0.0100	0.8500	0.0580	0.0214	96	-0.26	032.43	0
032	Potassium (%)	0142	0.8396	0.0260	0.8500	0.0580	0.0214	96	-0.18	032.31	0
032	Potassium (%)	0171	0.8400	0.0000	0.8500	0.0580	0.0214	96	-0.17	032.41	0
032	Potassium (%)	0939	0.8400	0.0000	0.8500	0.0580	0.0214	96	-0.17	032.53	0
032	Potassium (%)	0425	0.8400	0.0800	0.8500	0.0580	0.0214	96	-0.17	032.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
032	Potassium (%)	2246	0.8447	0.0032	0.8500	0.0580	0.0214	96	-0.09	032.33	0
032	Potassium (%)	0921	0.8450	0.0560	0.8500	0.0580	0.0214	96	-0.09	032.31	0
032	Potassium (%)	0226	0.8450	0.0100	0.8500	0.0580	0.0214	96	-0.09	032.41	0
032	Potassium (%)	0413	0.8450	0.0100	0.8500	0.0580	0.0214	96	-0.09	032.41	0
032	Potassium (%)	0910	0.8450	0.0700	0.8500	0.0580	0.0214	96	-0.09	032.41	0
032	Potassium (%)	0964	0.8450	0.0010	0.8500	0.0580	0.0214	96	-0.09	032.43	0
032	Potassium (%)	0553	0.8455	0.0150	0.8500	0.0580	0.0214	96	-0.08	032.53	0
032	Potassium (%)	0693	0.8464	0.0010	0.8500	0.0580	0.0214	96	-0.06	032.42	0
032	Potassium (%)	0914	0.8485	0.0531	0.8500	0.0580	0.0214	96	-0.03	032.43	0
032	Potassium (%)	0042	0.8495	0.0950	0.8500	0.0580	0.0214	96	-0.01	032.42	0
032	Potassium (%)	0169	0.8500	0.0400	0.8500	0.0580	0.0214	96	0.00	032.32	0
032	Potassium (%)	0139	0.8500	0.0000	0.8500	0.0580	0.0214	96	0.00	032.41	0
032	Potassium (%)	0035	0.8550	0.0100	0.8500	0.0580	0.0214	96	0.09	032.42	0
032	Potassium (%)	0278	0.8550	0.0100	0.8500	0.0580	0.0214	96	0.09	032.42	0
032	Potassium (%)	0345	0.8555	0.0110	0.8500	0.0580	0.0214	96	0.10	032.43	0
032	Potassium (%)	0884	0.8600	0.0000	0.8500	0.0580	0.0214	96	0.17	032.02	0
032	Potassium (%)	0948	0.8615	0.0210	0.8500	0.0580	0.0214	96	0.20	032.31	0
032	Potassium (%)	0512	0.8634	0.0371	0.8500	0.0580	0.0214	96	0.23	032.43	0
032	Potassium (%)	0038	0.8645	0.0430	0.8500	0.0580	0.0214	96	0.25	032.43	0
032	Potassium (%)	0100	0.8650	0.0300	0.8500	0.0580	0.0214	96	0.26	032.99	0
032	Potassium (%)	0918	0.8660	0.0260	0.8500	0.0580	0.0214	96	0.28	032.43	0
032	Potassium (%)	0653	0.8680	0.0080	0.8500	0.0580	0.0214	96	0.31	032.41	0
032	Potassium (%)	0692	0.8700	0.0400	0.8500	0.0580	0.0214	96	0.34	032.42	0
032	Potassium (%)	0682	0.8700	0.0000	0.8500	0.0580	0.0214	96	0.34	032.43	0
032	Potassium (%)	0043	0.8703	0.0064	0.8500	0.0580	0.0214	96	0.35	032.43	0
032	Potassium (%)	0870	0.8706	0.0065	0.8500	0.0580	0.0214	96	0.35	032.42	0
032	Potassium (%)	0047	0.8750	0.0300	0.8500	0.0580	0.0214	96	0.43	032.42	0
032	Potassium (%)	0010	0.8800	0.0200	0.8500	0.0580	0.0214	96	0.52	032.43	0
032	Potassium (%)	0610	0.8800	0.0000	0.8500	0.0580	0.0214	96	0.52	032.43	0
032	Potassium (%)	0035	0.8850	0.0100	0.8500	0.0580	0.0214	96	0.60	032.43	0
032	Potassium (%)	0407	0.8868	0.0988	0.8500	0.0580	0.0214	96	0.63	032.41	0
032	Potassium (%)	0148	0.8900	0.0200	0.8500	0.0580	0.0214	96	0.69	032.41	0
032	Potassium (%)	0675	0.8900	0.0200	0.8500	0.0580	0.0214	96	0.69	032.43	0
032	Potassium (%)	0560	0.8926	0.0684	0.8500	0.0580	0.0214	96	0.73	032.52	0
032	Potassium (%)	0199	0.8940	0.0100	0.8500	0.0580	0.0214	96	0.76	032.53	0
032	Potassium (%)	0066	0.8950	0.0160	0.8500	0.0580	0.0214	96	0.78	032.43	0
032	Potassium (%)	0297	0.8950	0.0300	0.8500	0.0580	0.0214	96	0.78	032.43	0
032	Potassium (%)	0037	0.8970	0.0040	0.8500	0.0580	0.0214	96	0.81	032.42	0
032	Potassium (%)	0027	0.8985	0.0050	0.8500	0.0580	0.0214	96	0.84	032.43	0
032	Potassium (%)	0726	0.9000	0.0000	0.8500	0.0580	0.0214	96	0.86	032.42	0
032	Potassium (%)	0357	0.9016	0.0069	0.8500	0.0580	0.0214	96	0.89	032.42	0
032	Potassium (%)	0510	0.9100	0.0200	0.8500	0.0580	0.0214	96	1.03	032.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
032	Potassium (%)	2089	0.9100	0.0000	0.8500	0.0580	0.0214	96	1.03	032.43	0
032	Potassium (%)	0202	0.9125	0.0630	0.8500	0.0580	0.0214	96	1.08	032.43	0
032	Potassium (%)	0032	0.9161	0.0099	0.8500	0.0580	0.0214	96	1.14	032.43	0
032	Potassium (%)	0353	0.9200	0.0600	0.8500	0.0580	0.0214	96	1.21	032.43	0
032	Potassium (%)	0190	0.9350	0.0100	0.8500	0.0580	0.0214	96	1.46	032.42	0
032	Potassium (%)	0017	0.9369	0.0159	0.8500	0.0580	0.0214	96	1.50	032.43	0
032	Potassium (%)	0407	0.9389	0.0024	0.8500	0.0580	0.0214	96	1.53	032.43	0
032	Potassium (%)	0089	0.9400	0.0000	0.8500	0.0580	0.0214	96	1.55	032.43	0
032	Potassium (%)	0968	0.9580	0.0140	0.8500	0.0580	0.0214	96	1.86	032.43	0
032	Potassium (%)	0263	0.9671	0.0150	0.8500	0.0580	0.0214	96	2.02	032.42	0
032	Potassium (%)	0009	0.9700	0.0600	0.8500	0.0580	0.0214	96	2.07	032.42	0
032	Potassium (%)	0520	0.9860	0.0160	0.8500	0.0580	0.0214	96	2.34	032.41	0
032	Potassium (%)	0366	1.015	0.0300	0.8500	0.0580	0.0214	96	2.84	032.42	0
032	Potassium (%)	0354	1.205	0.0100	0.8500	0.0580	0.0214	96	6.12	032.41	0
032	Potassium (%)	0003	0.7800	0.1200	0.8500	0.0580	0.0214	96	-1.21	032.41	1
032	Potassium (%)	2146	0.8750	0.1300	0.8500	0.0580	0.0214	96	0.43	032.51	1
033	Salt as chloride (%)	2129	1.029	0.0290	2.067	0.1011	0.0197	66	-10.27	033.99	0
033	Salt as chloride (%)	0536	1.395	0.0900	2.067	0.1011	0.0197	66	-6.64	033.99	0
033	Salt as chloride (%)	2246	1.500	0.0000	2.067	0.1011	0.0197	66	-5.60	033.00	0
033	Salt as chloride (%)	0875	1.630	0.0400	2.067	0.1011	0.0197	66	-4.32	033.99	0
033	Salt as chloride (%)	0693	1.780	0.0374	2.067	0.1011	0.0197	66	-2.83	033.00	0
033	Salt as chloride (%)	0265	1.865	0.0100	2.067	0.1011	0.0197	66	-1.99	033.03	0
033	Salt as chloride (%)	0017	1.870	0.0416	2.067	0.1011	0.0197	66	-1.95	033.99	0
033	Salt as chloride (%)	0032	1.887	0.0340	2.067	0.1011	0.0197	66	-1.78	033.99	0
033	Salt as chloride (%)	0893	1.915	0.0300	2.067	0.1011	0.0197	66	-1.50	033.00	0
033	Salt as chloride (%)	0190	1.920	0.0000	2.067	0.1011	0.0197	66	-1.45	033.03	0
033	Salt as chloride (%)	0726	1.920	0.0000	2.067	0.1011	0.0197	66	-1.45	033.03	0
033	Salt as chloride (%)	2295	1.955	0.0100	2.067	0.1011	0.0197	66	-1.10	033.00	0
033	Salt as chloride (%)	0505	1.960	0.0000	2.067	0.1011	0.0197	66	-1.06	033.03	0
033	Salt as chloride (%)	2076	1.970	0.0000	2.067	0.1011	0.0197	66	-0.96	033.00	0
033	Salt as chloride (%)	0619	1.980	0.0200	2.067	0.1011	0.0197	66	-0.86	033.01	0
033	Salt as chloride (%)	0353	1.995	0.0100	2.067	0.1011	0.0197	66	-0.71	033.00	0
033	Salt as chloride (%)	0003	2.000	0.0000	2.067	0.1011	0.0197	66	-0.66	033.01	0
033	Salt as chloride (%)	0539	2.010	0.0200	2.067	0.1011	0.0197	66	-0.56	033.00	0
033	Salt as chloride (%)	2073	2.010	0.0200	2.067	0.1011	0.0197	66	-0.56	033.99	0
033	Salt as chloride (%)	2181	2.015	0.0300	2.067	0.1011	0.0197	66	-0.51	033.99	0
033	Salt as chloride (%)	0171	2.020	0.0000	2.067	0.1011	0.0197	66	-0.46	033.00	0
033	Salt as chloride (%)	2193	2.035	0.0300	2.067	0.1011	0.0197	66	-0.31	033.00	0
033	Salt as chloride (%)	0682	2.050	0.0000	2.067	0.1011	0.0197	66	-0.17	033.99	0
033	Salt as chloride (%)	2009	2.050	0.0080	2.067	0.1011	0.0197	66	-0.16	033.00	0
033	Salt as chloride (%)	2022	2.055	0.0300	2.067	0.1011	0.0197	66	-0.12	033.00	0
033	Salt as chloride (%)	0884	2.060	0.0200	2.067	0.1011	0.0197	66	-0.07	033.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
033	Salt as chloride (%)	0083	2.065	0.0100	2.067	0.1011	0.0197	66	-0.02	033.01	0
033	Salt as chloride (%)	0413	2.070	0.0200	2.067	0.1011	0.0197	66	0.03	033.01	0
033	Salt as chloride (%)	0878	2.080	0.0600	2.067	0.1011	0.0197	66	0.13	033.00	0
033	Salt as chloride (%)	0164	2.080	0.0200	2.067	0.1011	0.0197	66	0.13	033.01	0
033	Salt as chloride (%)	0123	2.090	0.0200	2.067	0.1011	0.0197	66	0.23	033.00	0
033	Salt as chloride (%)	2109	2.095	0.0080	2.067	0.1011	0.0197	66	0.28	033.01	0
033	Salt as chloride (%)	0309	2.098	0.0166	2.067	0.1011	0.0197	66	0.31	033.00	0
033	Salt as chloride (%)	0144	2.100	0.0000	2.067	0.1011	0.0197	66	0.33	033.03	0
033	Salt as chloride (%)	0868	2.100	0.0400	2.067	0.1011	0.0197	66	0.33	033.05	0
033	Salt as chloride (%)	0001	2.110	0.0000	2.067	0.1011	0.0197	66	0.43	033.01	0
033	Salt as chloride (%)	0098	2.110	0.0000	2.067	0.1011	0.0197	66	0.43	033.01	0
033	Salt as chloride (%)	0298	2.110	0.0200	2.067	0.1011	0.0197	66	0.43	033.00	0
033	Salt as chloride (%)	0407	2.115	0.0270	2.067	0.1011	0.0197	66	0.47	033.01	0
033	Salt as chloride (%)	0169	2.115	0.0300	2.067	0.1011	0.0197	66	0.48	033.00	0
033	Salt as chloride (%)	0511	2.115	0.0300	2.067	0.1011	0.0197	66	0.48	033.00	0
033	Salt as chloride (%)	0004	2.115	0.0500	2.067	0.1011	0.0197	66	0.48	033.01	0
033	Salt as chloride (%)	2146	2.115	0.0100	2.067	0.1011	0.0197	66	0.48	033.01	0
033	Salt as chloride (%)	0229	2.119	0.0210	2.067	0.1011	0.0197	66	0.51	033.01	0
033	Salt as chloride (%)	0366	2.120	0.0000	2.067	0.1011	0.0197	66	0.53	033.00	0
033	Salt as chloride (%)	2196	2.120	0.0000	2.067	0.1011	0.0197	66	0.53	033.00	0
033	Salt as chloride (%)	0897	2.125	0.0100	2.067	0.1011	0.0197	66	0.58	033.00	0
033	Salt as chloride (%)	0590	2.125	0.0300	2.067	0.1011	0.0197	66	0.58	033.01	0
033	Salt as chloride (%)	0148	2.130	0.0200	2.067	0.1011	0.0197	66	0.63	033.01	0
033	Salt as chloride (%)	0242	2.130	0.0000	2.067	0.1011	0.0197	66	0.63	033.01	0
033	Salt as chloride (%)	0948	2.130	0.0400	2.067	0.1011	0.0197	66	0.63	033.01	0
033	Salt as chloride (%)	0175	2.135	0.0300	2.067	0.1011	0.0197	66	0.68	033.01	0
033	Salt as chloride (%)	0689	2.135	0.0300	2.067	0.1011	0.0197	66	0.68	033.05	0
033	Salt as chloride (%)	0510	2.144	0.0000	2.067	0.1011	0.0197	66	0.76	033.01	0
033	Salt as chloride (%)	0354	2.150	0.0200	2.067	0.1011	0.0197	66	0.82	033.01	0
033	Salt as chloride (%)	0100	2.155	0.0100	2.067	0.1011	0.0197	66	0.87	033.01	0
033	Salt as chloride (%)	0199	2.155	0.0100	2.067	0.1011	0.0197	66	0.87	033.01	0
033	Salt as chloride (%)	0650	2.160	0.0200	2.067	0.1011	0.0197	66	0.92	033.99	0
033	Salt as chloride (%)	2062	2.160	0.0022	2.067	0.1011	0.0197	66	0.93	033.00	0
033	Salt as chloride (%)	0226	2.165	0.0300	2.067	0.1011	0.0197	66	0.97	033.01	0
033	Salt as chloride (%)	0939	2.165	0.0500	2.067	0.1011	0.0197	66	0.97	033.01	0
033	Salt as chloride (%)	0019	2.170	0.0200	2.067	0.1011	0.0197	66	1.02	033.01	0
033	Salt as chloride (%)	0921	2.180	0.0000	2.067	0.1011	0.0197	66	1.12	033.00	0
033	Salt as chloride (%)	0881	2.191	0.0230	2.067	0.1011	0.0197	66	1.22	033.00	0
033	Salt as chloride (%)	0675	2.200	0.0400	2.067	0.1011	0.0197	66	1.32	033.00	0
033	Salt as chloride (%)	0278	2.440	0.0200	2.067	0.1011	0.0197	66	3.69	033.01	0
033	Salt as chloride (%)	0910	1.380	0.1800	2.067	0.1011	0.0197	66	-6.79	033.99	1
033	Salt as chloride (%)	0358	1.845	0.1500	2.067	0.1011	0.0197	66	-2.19	033.99	1

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
034	Selenium (ppm)	0682	1.240	0.0000	2.004	0.5327	0.1513	26	-1.43	034.43	0
034	Selenium (ppm)	0563	1.304	0.3376	2.004	0.5327	0.1513	26	-1.31	034.04	0
034	Selenium (ppm)	0171	1.347	0.5070	2.004	0.5327	0.1513	26	-1.23	034.04	0
034	Selenium (ppm)	0619	1.425	0.0300	2.004	0.5327	0.1513	26	-1.09	034.41	0
034	Selenium (ppm)	0199	1.450	0.0200	2.004	0.5327	0.1513	26	-1.04	034.53	0
034	Selenium (ppm)	0032	1.520	0.0200	2.004	0.5327	0.1513	26	-0.91	034.53	0
034	Selenium (ppm)	0148	1.792	0.0520	2.004	0.5327	0.1513	26	-0.40	034.41	0
034	Selenium (ppm)	0512	1.816	0.1910	2.004	0.5327	0.1513	26	-0.35	034.04	0
034	Selenium (ppm)	0692	1.850	0.1000	2.004	0.5327	0.1513	26	-0.29	034.42	0
034	Selenium (ppm)	0553	1.865	0.1500	2.004	0.5327	0.1513	26	-0.26	034.53	0
034	Selenium (ppm)	0164	1.885	0.0700	2.004	0.5327	0.1513	26	-0.22	034.53	0
034	Selenium (ppm)	0169	1.975	0.1300	2.004	0.5327	0.1513	26	-0.05	034.04	0
034	Selenium (ppm)	0918	2.004	0.0051	2.004	0.5327	0.1513	26	0.00	034.53	0
034	Selenium (ppm)	0038	2.023	0.1204	2.004	0.5327	0.1513	26	0.03	034.01	0
034	Selenium (ppm)	0047	2.035	0.0100	2.004	0.5327	0.1513	26	0.06	034.52	0
034	Selenium (ppm)	0610	2.035	0.1100	2.004	0.5327	0.1513	26	0.06	034.53	0
034	Selenium (ppm)	0939	2.035	0.3700	2.004	0.5327	0.1513	26	0.06	034.53	0
034	Selenium (ppm)	0003	2.058	0.0540	2.004	0.5327	0.1513	26	0.10	034.53	0
034	Selenium (ppm)	0560	2.129	0.0330	2.004	0.5327	0.1513	26	0.23	034.53	0
034	Selenium (ppm)	0914	2.238	0.2550	2.004	0.5327	0.1513	26	0.44	034.53	0
034	Selenium (ppm)	0910	2.310	0.0200	2.004	0.5327	0.1513	26	0.57	034.52	0
034	Selenium (ppm)	0968	2.676	0.0390	2.004	0.5327	0.1513	26	1.26	034.43	0
034	Selenium (ppm)	2012	2.704	0.0590	2.004	0.5327	0.1513	26	1.31	034.53	0
034	Selenium (ppm)	0407	2.785	0.5100	2.004	0.5327	0.1513	26	1.47	034.41	0
034	Selenium (ppm)	0407	2.910	0.4400	2.004	0.5327	0.1513	26	1.70	034.43	0
034	Selenium (ppm)	0190	3.470	0.3000	2.004	0.5327	0.1513	26	2.75	034.99	0
034	Selenium (ppm)	0066	231.6	16.51	2.004	0.5327	0.1513	26	430.94	034.43	2
034	Selenium (ppm)	2302	< 0.01		2.004	0.5327	0.1513	26		034.99	5
035	Sodium (%)	2129	0.5352	0.0001	0.7046	0.0415	0.0186	85	-4.08	035.42	0
035	Sodium (%)	2259	0.6065	0.0150	0.7046	0.0415	0.0186	85	-2.36	035.41	0
035	Sodium (%)	0598	0.6142	0.0109	0.7046	0.0415	0.0186	85	-2.18	035.41	0
035	Sodium (%)	0529	0.6210	0.0020	0.7046	0.0415	0.0186	85	-2.01	035.31	0
035	Sodium (%)	0939	0.6350	0.0100	0.7046	0.0415	0.0186	85	-1.68	035.53	0
035	Sodium (%)	0298	0.6400	0.0200	0.7046	0.0415	0.0186	85	-1.56	035.41	0
035	Sodium (%)	2012	0.6450	0.0100	0.7046	0.0415	0.0186	85	-1.44	035.41	0
035	Sodium (%)	0622	0.6477	0.0089	0.7046	0.0415	0.0186	85	-1.37	035.31	0
035	Sodium (%)	0620	0.6478	0.0452	0.7046	0.0415	0.0186	85	-1.37	035.31	0
035	Sodium (%)	0171	0.6500	0.0000	0.7046	0.0415	0.0186	85	-1.32	035.41	0
035	Sodium (%)	0265	0.6550	0.0100	0.7046	0.0415	0.0186	85	-1.19	035.42	0
035	Sodium (%)	0505	0.6550	0.0500	0.7046	0.0415	0.0186	85	-1.19	035.43	0
035	Sodium (%)	0229	0.6565	0.0370	0.7046	0.0415	0.0186	85	-1.16	035.41	0
035	Sodium (%)	0868	0.6635	0.0110	0.7046	0.0415	0.0186	85	-0.99	035.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
035	Sodium (%)	0868	0.6650	0.0100	0.7046	0.0415	0.0186	85	-0.95	035.31	0
035	Sodium (%)	2146	0.6650	0.0100	0.7046	0.0415	0.0186	85	-0.95	035.31	0
035	Sodium (%)	0354	0.6650	0.0100	0.7046	0.0415	0.0186	85	-0.95	035.41	0
035	Sodium (%)	0910	0.6700	0.0600	0.7046	0.0415	0.0186	85	-0.83	035.41	0
035	Sodium (%)	0144	0.6700	0.0000	0.7046	0.0415	0.0186	85	-0.83	035.41	0
035	Sodium (%)	0042	0.6745	0.0630	0.7046	0.0415	0.0186	85	-0.73	035.42	0
035	Sodium (%)	0884	0.6750	0.0300	0.7046	0.0415	0.0186	85	-0.71	035.02	0
035	Sodium (%)	0619	0.6775	0.0090	0.7046	0.0415	0.0186	85	-0.65	035.41	0
035	Sodium (%)	0019	0.6800	0.0000	0.7046	0.0415	0.0186	85	-0.59	035.41	0
035	Sodium (%)	0089	0.6800	0.0000	0.7046	0.0415	0.0186	85	-0.59	035.43	0
035	Sodium (%)	0098	0.6800	0.0000	0.7046	0.0415	0.0186	85	-0.59	035.43	0
035	Sodium (%)	0939	0.6850	0.0100	0.7046	0.0415	0.0186	85	-0.47	035.31	0
035	Sodium (%)	0242	0.6850	0.0100	0.7046	0.0415	0.0186	85	-0.47	035.99	0
035	Sodium (%)	0142	0.6855	0.0305	0.7046	0.0415	0.0186	85	-0.46	035.31	0
035	Sodium (%)	0366	0.6880	0.0360	0.7046	0.0415	0.0186	85	-0.40	035.42	0
035	Sodium (%)	0098	0.6900	0.0400	0.7046	0.0415	0.0186	85	-0.35	035.41	0
035	Sodium (%)	0878	0.6950	0.0100	0.7046	0.0415	0.0186	85	-0.23	035.41	0
035	Sodium (%)	0345	0.6960	0.0080	0.7046	0.0415	0.0186	85	-0.21	035.43	0
035	Sodium (%)	0199	0.6970	0.0180	0.7046	0.0415	0.0186	85	-0.18	035.53	0
035	Sodium (%)	0693	0.6983	0.0087	0.7046	0.0415	0.0186	85	-0.15	035.42	0
035	Sodium (%)	0148	0.7000	0.0200	0.7046	0.0415	0.0186	85	-0.11	035.41	0
035	Sodium (%)	0413	0.7000	0.0400	0.7046	0.0415	0.0186	85	-0.11	035.41	0
035	Sodium (%)	0297	0.7000	0.0400	0.7046	0.0415	0.0186	85	-0.11	035.43	0
035	Sodium (%)	0510	0.7040	0.0040	0.7046	0.0415	0.0186	85	-0.01	035.43	0
035	Sodium (%)	0154	0.7042	0.0027	0.7046	0.0415	0.0186	85	-0.01	035.52	0
035	Sodium (%)	0164	0.7045	0.0310	0.7046	0.0415	0.0186	85	0.00	035.41	0
035	Sodium (%)	0083	0.7050	0.0100	0.7046	0.0415	0.0186	85	0.01	035.43	0
035	Sodium (%)	0653	0.7070	0.0020	0.7046	0.0415	0.0186	85	0.06	035.41	0
035	Sodium (%)	0512	0.7073	0.0519	0.7046	0.0415	0.0186	85	0.06	035.43	0
035	Sodium (%)	0553	0.7075	0.0170	0.7046	0.0415	0.0186	85	0.07	035.53	0
035	Sodium (%)	0563	0.7093	0.0029	0.7046	0.0415	0.0186	85	0.11	035.31	0
035	Sodium (%)	2196	0.7100	0.0000	0.7046	0.0415	0.0186	85	0.13	035.31	0
035	Sodium (%)	0123	0.7100	0.0200	0.7046	0.0415	0.0186	85	0.13	035.41	0
035	Sodium (%)	0035	0.7100	0.0000	0.7046	0.0415	0.0186	85	0.13	035.42	0
035	Sodium (%)	0035	0.7100	0.0000	0.7046	0.0415	0.0186	85	0.13	035.43	0
035	Sodium (%)	0682	0.7100	0.0000	0.7046	0.0415	0.0186	85	0.13	035.43	0
035	Sodium (%)	0590	0.7100	0.0000	0.7046	0.0415	0.0186	85	0.13	035.99	0
035	Sodium (%)	0948	0.7150	0.0180	0.7046	0.0415	0.0186	85	0.25	035.31	0
035	Sodium (%)	0037	0.7180	0.0040	0.7046	0.0415	0.0186	85	0.32	035.42	0
035	Sodium (%)	0353	0.7200	0.0200	0.7046	0.0415	0.0186	85	0.37	035.43	0
035	Sodium (%)	0918	0.7205	0.0150	0.7046	0.0415	0.0186	85	0.38	035.43	0
035	Sodium (%)	0914	0.7250	0.0681	0.7046	0.0415	0.0186	85	0.49	035.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
035	Sodium (%)	2146	0.7250	0.0100	0.7046	0.0415	0.0186	85	0.49	035.51	0
035	Sodium (%)	0870	0.7271	0.0228	0.7046	0.0415	0.0186	85	0.54	035.42	0
035	Sodium (%)	0139	0.7285	0.0170	0.7046	0.0415	0.0186	85	0.58	035.41	0
035	Sodium (%)	0006	0.7285	0.0130	0.7046	0.0415	0.0186	85	0.58	035.43	0
035	Sodium (%)	0964	0.7294	0.0463	0.7046	0.0415	0.0186	85	0.60	035.43	0
035	Sodium (%)	0226	0.7300	0.0200	0.7046	0.0415	0.0186	85	0.61	035.41	0
035	Sodium (%)	0358	0.7300	0.0600	0.7046	0.0415	0.0186	85	0.61	035.41	0
035	Sodium (%)	0278	0.7300	0.0200	0.7046	0.0415	0.0186	85	0.61	035.42	0
035	Sodium (%)	0675	0.7300	0.0200	0.7046	0.0415	0.0186	85	0.61	035.43	0
035	Sodium (%)	0572	0.7305	0.0030	0.7046	0.0415	0.0186	85	0.62	035.53	0
035	Sodium (%)	2246	0.7318	0.0011	0.7046	0.0415	0.0186	85	0.65	035.33	0
035	Sodium (%)	0169	0.7350	0.0300	0.7046	0.0415	0.0186	85	0.73	035.32	0
035	Sodium (%)	0610	0.7350	0.0300	0.7046	0.0415	0.0186	85	0.73	035.43	0
035	Sodium (%)	0017	0.7356	0.0164	0.7046	0.0415	0.0186	85	0.75	035.43	0
035	Sodium (%)	2089	0.7400	0.0000	0.7046	0.0415	0.0186	85	0.85	035.43	0
035	Sodium (%)	0881	0.7420	0.0040	0.7046	0.0415	0.0186	85	0.90	035.43	0
035	Sodium (%)	0692	0.7500	0.0400	0.7046	0.0415	0.0186	85	1.09	035.42	0
035	Sodium (%)	0100	0.7500	0.0200	0.7046	0.0415	0.0186	85	1.09	035.99	0
035	Sodium (%)	0202	0.7525	0.0650	0.7046	0.0415	0.0186	85	1.15	035.43	0
035	Sodium (%)	0407	0.7533	0.0256	0.7046	0.0415	0.0186	85	1.17	035.41	0
035	Sodium (%)	0425	0.7550	0.0700	0.7046	0.0415	0.0186	85	1.21	035.43	0
035	Sodium (%)	0263	0.7576	0.0028	0.7046	0.0415	0.0186	85	1.28	035.42	0
035	Sodium (%)	0726	0.7600	0.0000	0.7046	0.0415	0.0186	85	1.33	035.42	0
035	Sodium (%)	0294	0.7700	0.0000	0.7046	0.0415	0.0186	85	1.58	035.42	0
035	Sodium (%)	0407	0.7846	0.0320	0.7046	0.0415	0.0186	85	1.93	035.43	0
035	Sodium (%)	0536	0.8200	0.0200	0.7046	0.0415	0.0186	85	2.78	035.31	0
035	Sodium (%)	0520	0.8271	0.0003	0.7046	0.0415	0.0186	85	2.95	035.41	0
035	Sodium (%)	0038	0.8880	0.0120	0.7046	0.0415	0.0186	85	4.42	035.43	0
035	Sodium (%)	0066	0.9420	0.0160	0.7046	0.0415	0.0186	85	5.72	035.43	0
035	Sodium (%)	0049	0.6700	0.1000	0.7046	0.0415	0.0186	85	-0.83	035.41	1
035	Sodium (%)	0560	0.8066	0.0983	0.7046	0.0415	0.0186	85	2.46	035.52	1
035	Sodium (%)	0889	0.1050	0.0000	0.7046	0.0415	0.0186	85	-14.44	035.99	2
036	Sulfur (%)	0889	0.2900	0.0100	0.3842	0.0262	0.0121	40	-3.59	036.99	0
036	Sulfur (%)	0229	0.3400	0.0000	0.3842	0.0262	0.0121	40	-1.68	036.42	0
036	Sulfur (%)	0870	0.3504	0.0124	0.3842	0.0262	0.0121	40	-1.29	036.42	0
036	Sulfur (%)	0693	0.3535	0.0058	0.3842	0.0262	0.0121	40	-1.17	036.42	0
036	Sulfur (%)	0407	0.3542	0.0152	0.3842	0.0262	0.0121	40	-1.14	036.43	0
036	Sulfur (%)	0297	0.3550	0.0100	0.3842	0.0262	0.0121	40	-1.11	036.43	0
036	Sulfur (%)	0366	0.3600	0.0400	0.3842	0.0262	0.0121	40	-0.92	036.42	0
036	Sulfur (%)	0035	0.3650	0.0100	0.3842	0.0262	0.0121	40	-0.73	036.43	0
036	Sulfur (%)	0083	0.3650	0.0100	0.3842	0.0262	0.0121	40	-0.73	036.43	0
036	Sulfur (%)	0169	0.3650	0.0100	0.3842	0.0262	0.0121	40	-0.73	036.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
036	Sulfur (%)	0357	0.3674	0.0094	0.3842	0.0262	0.0121	40	-0.64	036.42	0
036	Sulfur (%)	0229	0.3680	0.0060	0.3842	0.0262	0.0121	40	-0.62	036.04	0
036	Sulfur (%)	0242	0.3700	0.0000	0.3842	0.0262	0.0121	40	-0.54	036.99	0
036	Sulfur (%)	0918	0.3740	0.0060	0.3842	0.0262	0.0121	40	-0.39	036.43	0
036	Sulfur (%)	2053	0.3750	0.0100	0.3842	0.0262	0.0121	40	-0.35	036.01	0
036	Sulfur (%)	0265	0.3750	0.0100	0.3842	0.0262	0.0121	40	-0.35	036.42	0
036	Sulfur (%)	0038	0.3762	0.0025	0.3842	0.0262	0.0121	40	-0.31	036.00	0
036	Sulfur (%)	2129	0.3783	0.0165	0.3842	0.0262	0.0121	40	-0.23	036.42	0
036	Sulfur (%)	0692	0.3800	0.0400	0.3842	0.0262	0.0121	40	-0.16	036.42	0
036	Sulfur (%)	0353	0.3800	0.0200	0.3842	0.0262	0.0121	40	-0.16	036.43	0
036	Sulfur (%)	0164	0.3850	0.0100	0.3842	0.0262	0.0121	40	0.03	036.42	0
036	Sulfur (%)	0278	0.3850	0.0100	0.3842	0.0262	0.0121	40	0.03	036.42	0
036	Sulfur (%)	0294	0.3850	0.0100	0.3842	0.0262	0.0121	40	0.03	036.42	0
036	Sulfur (%)	0345	0.3885	0.0070	0.3842	0.0262	0.0121	40	0.16	036.43	0
036	Sulfur (%)	0726	0.3900	0.0000	0.3842	0.0262	0.0121	40	0.22	036.42	0
036	Sulfur (%)	0148	0.3910	0.0120	0.3842	0.0262	0.0121	40	0.26	036.04	0
036	Sulfur (%)	0910	0.3913	0.0226	0.3842	0.0262	0.0121	40	0.27	036.52	0
036	Sulfur (%)	0553	0.3940	0.0140	0.3842	0.0262	0.0121	40	0.37	036.53	0
036	Sulfur (%)	0098	0.3993	0.0096	0.3842	0.0262	0.0121	40	0.58	036.43	0
036	Sulfur (%)	0037	0.4025	0.0070	0.3842	0.0262	0.0121	40	0.70	036.42	0
036	Sulfur (%)	0226	0.4050	0.0100	0.3842	0.0262	0.0121	40	0.79	036.04	0
036	Sulfur (%)	0202	0.4055	0.0370	0.3842	0.0262	0.0121	40	0.81	036.43	0
036	Sulfur (%)	0708	0.4080	0.0060	0.3842	0.0262	0.0121	40	0.91	036.42	0
036	Sulfur (%)	0038	0.4080	0.0040	0.3842	0.0262	0.0121	40	0.91	036.43	0
036	Sulfur (%)	0560	0.4181	0.0113	0.3842	0.0262	0.0121	40	1.29	036.52	0
036	Sulfur (%)	0032	0.4187	0.0091	0.3842	0.0262	0.0121	40	1.31	036.43	0
036	Sulfur (%)	2089	0.4200	0.0400	0.3842	0.0262	0.0121	40	1.37	036.43	0
036	Sulfur (%)	0510	0.4300	0.0000	0.3842	0.0262	0.0121	40	1.75	036.43	0
036	Sulfur (%)	0171	0.4350	0.0100	0.3842	0.0262	0.0121	40	1.94	036.42	0
036	Sulfur (%)	0505	0.4350	0.0100	0.3842	0.0262	0.0121	40	1.94	036.43	0
036	Sulfur (%)	0042	0.3805	0.0630	0.3842	0.0262	0.0121	40	-0.14	036.42	1
037	Zinc (ppm)	2129	153.7	1.100	493.7	69.18	18.02	91	-4.92	037.42	0
037	Zinc (ppm)	0199	332.2	22.31	493.7	69.18	18.02	91	-2.33	037.53	0
037	Zinc (ppm)	2012	344.5	11.69	493.7	69.18	18.02	91	-2.16	037.41	0
037	Zinc (ppm)	0622	369.5	3.000	493.7	69.18	18.02	91	-1.79	037.31	0
037	Zinc (ppm)	0536	372.3	11.70	493.7	69.18	18.02	91	-1.75	037.31	0
037	Zinc (ppm)	0049	372.3	21.38	493.7	69.18	18.02	91	-1.75	037.41	0
037	Zinc (ppm)	0004	377.4	56.93	493.7	69.18	18.02	91	-1.68	037.41	0
037	Zinc (ppm)	0619	379.5	5.000	493.7	69.18	18.02	91	-1.65	037.41	0
037	Zinc (ppm)	0889	389.3	11.71	493.7	69.18	18.02	91	-1.51	037.99	0
037	Zinc (ppm)	0003	394.0	22.00	493.7	69.18	18.02	91	-1.44	037.41	0
037	Zinc (ppm)	0074	403.5	81.00	493.7	69.18	18.02	91	-1.30	037.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
037	Zinc (ppm)	0164	409.5	5.000	493.7	69.18	18.02	91	-1.22	037.44	0
037	Zinc (ppm)	0653	411.8	2.300	493.7	69.18	18.02	91	-1.18	037.41	0
037	Zinc (ppm)	2144	422.1	14.63	493.7	69.18	18.02	91	-1.03	037.99	0
037	Zinc (ppm)	0868	425.0	4.000	493.7	69.18	18.02	91	-0.99	037.31	0
037	Zinc (ppm)	0354	427.1	9.490	493.7	69.18	18.02	91	-0.96	037.41	0
037	Zinc (ppm)	0171	436.6	9.450	493.7	69.18	18.02	91	-0.82	037.41	0
037	Zinc (ppm)	0047	439.0	30.00	493.7	69.18	18.02	91	-0.79	037.52	0
037	Zinc (ppm)	0098	439.7	24.80	493.7	69.18	18.02	91	-0.78	037.41	0
037	Zinc (ppm)	0175	447.0	30.00	493.7	69.18	18.02	91	-0.67	037.31	0
037	Zinc (ppm)	2196	447.6	0.0000	493.7	69.18	18.02	91	-0.67	037.31	0
037	Zinc (ppm)	0598	447.9	29.80	493.7	69.18	18.02	91	-0.66	037.41	0
037	Zinc (ppm)	0590	450.0	20.00	493.7	69.18	18.02	91	-0.63	037.99	0
037	Zinc (ppm)	0019	450.4	39.40	493.7	69.18	18.02	91	-0.63	037.41	0
037	Zinc (ppm)	0914	451.8	50.20	493.7	69.18	18.02	91	-0.61	037.43	0
037	Zinc (ppm)	0910	453.5	41.00	493.7	69.18	18.02	91	-0.58	037.41	0
037	Zinc (ppm)	0229	457.3	22.00	493.7	69.18	18.02	91	-0.53	037.41	0
037	Zinc (ppm)	0001	458.2	9.300	493.7	69.18	18.02	91	-0.51	037.31	0
037	Zinc (ppm)	0529	459.8	46.90	493.7	69.18	18.02	91	-0.49	037.31	0
037	Zinc (ppm)	0878	460.0	14.00	493.7	69.18	18.02	91	-0.49	037.41	0
037	Zinc (ppm)	0294	462.5	12.48	493.7	69.18	18.02	91	-0.45	037.42	0
037	Zinc (ppm)	0226	463.0	0.0800	493.7	69.18	18.02	91	-0.44	037.41	0
037	Zinc (ppm)	0358	466.4	12.16	493.7	69.18	18.02	91	-0.39	037.41	0
037	Zinc (ppm)	0620	468.8	40.11	493.7	69.18	18.02	91	-0.36	037.31	0
037	Zinc (ppm)	0563	473.2	9.437	493.7	69.18	18.02	91	-0.30	037.31	0
037	Zinc (ppm)	0038	474.8	9.500	493.7	69.18	18.02	91	-0.27	037.32	0
037	Zinc (ppm)	0939	481.0	4.000	493.7	69.18	18.02	91	-0.18	037.53	0
037	Zinc (ppm)	0693	482.0	8.598	493.7	69.18	18.02	91	-0.17	037.42	0
037	Zinc (ppm)	0353	483.0	8.000	493.7	69.18	18.02	91	-0.15	037.43	0
037	Zinc (ppm)	0875	484.5	5.000	493.7	69.18	18.02	91	-0.13	037.43	0
037	Zinc (ppm)	0066	484.7	25.08	493.7	69.18	18.02	91	-0.13	037.43	0
037	Zinc (ppm)	0345	492.4	3.700	493.7	69.18	18.02	91	-0.02	037.43	0
037	Zinc (ppm)	0265	494.5	59.00	493.7	69.18	18.02	91	0.01	037.42	0
037	Zinc (ppm)	0357	494.8	3.563	493.7	69.18	18.02	91	0.02	037.42	0
037	Zinc (ppm)	0425	496.2	20.70	493.7	69.18	18.02	91	0.04	037.43	0
037	Zinc (ppm)	0169	498.0	16.00	493.7	69.18	18.02	91	0.06	037.43	0
037	Zinc (ppm)	0083	498.5	9.000	493.7	69.18	18.02	91	0.07	037.43	0
037	Zinc (ppm)	0572	499.0	20.00	493.7	69.18	18.02	91	0.08	037.53	0
037	Zinc (ppm)	0918	499.3	7.220	493.7	69.18	18.02	91	0.08	037.43	0
037	Zinc (ppm)	0610	501.5	43.00	493.7	69.18	18.02	91	0.11	037.43	0
037	Zinc (ppm)	0968	506.0	0.0000	493.7	69.18	18.02	91	0.18	037.43	0
037	Zinc (ppm)	0242	506.5	7.000	493.7	69.18	18.02	91	0.19	037.99	0
037	Zinc (ppm)	0366	508.0	6.000	493.7	69.18	18.02	91	0.21	037.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
037	Zinc (ppm)	0692	513.0	32.00	493.7	69.18	18.02	91	0.28	037.42	0
037	Zinc (ppm)	0520	517.5	1.000	493.7	69.18	18.02	91	0.34	037.41	0
037	Zinc (ppm)	0884	523.0	23.80	493.7	69.18	18.02	91	0.42	037.31	0
037	Zinc (ppm)	0505	523.1	23.50	493.7	69.18	18.02	91	0.42	037.43	0
037	Zinc (ppm)	0675	523.9	10.42	493.7	69.18	18.02	91	0.44	037.43	0
037	Zinc (ppm)	0017	524.8	14.58	493.7	69.18	18.02	91	0.45	037.43	0
037	Zinc (ppm)	0263	525.9	2.980	493.7	69.18	18.02	91	0.47	037.42	0
037	Zinc (ppm)	0848	526.5	4.650	493.7	69.18	18.02	91	0.47	037.41	0
037	Zinc (ppm)	0553	527.0	8.000	493.7	69.18	18.02	91	0.48	037.53	0
037	Zinc (ppm)	0964	529.2	17.91	493.7	69.18	18.02	91	0.51	037.43	0
037	Zinc (ppm)	2089	529.3	2.220	493.7	69.18	18.02	91	0.52	037.43	0
037	Zinc (ppm)	0043	531.3	0.6000	493.7	69.18	18.02	91	0.54	037.43	0
037	Zinc (ppm)	0278	533.3	7.500	493.7	69.18	18.02	91	0.57	037.42	0
037	Zinc (ppm)	0939	534.5	7.000	493.7	69.18	18.02	91	0.59	037.31	0
037	Zinc (ppm)	0009	539.0	8.000	493.7	69.18	18.02	91	0.66	037.42	0
037	Zinc (ppm)	0413	540.0	40.00	493.7	69.18	18.02	91	0.67	037.42	0
037	Zinc (ppm)	0190	545.1	4.830	493.7	69.18	18.02	91	0.74	037.42	0
037	Zinc (ppm)	0148	548.4	69.22	493.7	69.18	18.02	91	0.79	037.41	0
037	Zinc (ppm)	0870	549.3	4.700	493.7	69.18	18.02	91	0.80	037.42	0
037	Zinc (ppm)	0027	550.4	35.77	493.7	69.18	18.02	91	0.82	037.43	0
037	Zinc (ppm)	0035	552.5	37.00	493.7	69.18	18.02	91	0.85	037.43	0
037	Zinc (ppm)	0407	553.2	57.10	493.7	69.18	18.02	91	0.86	037.41	0
037	Zinc (ppm)	0202	553.8	4.540	493.7	69.18	18.02	91	0.87	037.43	0
037	Zinc (ppm)	2192	554.7	13.14	493.7	69.18	18.02	91	0.88	037.43	0
037	Zinc (ppm)	0297	559.0	30.00	493.7	69.18	18.02	91	0.94	037.43	0
037	Zinc (ppm)	0010	560.0	40.00	493.7	69.18	18.02	91	0.96	037.33	0
037	Zinc (ppm)	0098	563.1	22.90	493.7	69.18	18.02	91	1.00	037.43	0
037	Zinc (ppm)	0510	572.5	3.000	493.7	69.18	18.02	91	1.14	037.43	0
037	Zinc (ppm)	0512	574.9	13.60	493.7	69.18	18.02	91	1.17	037.43	0
037	Zinc (ppm)	0035	576.0	10.00	493.7	69.18	18.02	91	1.19	037.42	0
037	Zinc (ppm)	0726	578.7	2.800	493.7	69.18	18.02	91	1.23	037.42	0
037	Zinc (ppm)	0682	589.2	0.0000	493.7	69.18	18.02	91	1.38	037.43	0
037	Zinc (ppm)	0100	595.0	12.00	493.7	69.18	18.02	91	1.46	037.99	0
037	Zinc (ppm)	0038	600.0	12.00	493.7	69.18	18.02	91	1.54	037.43	0
037	Zinc (ppm)	0032	618.1	21.00	493.7	69.18	18.02	91	1.80	037.43	0
037	Zinc (ppm)	0037	625.0	29.10	493.7	69.18	18.02	91	1.90	037.42	0
037	Zinc (ppm)	0407	627.0	4.329	493.7	69.18	18.02	91	1.93	037.43	0
037	Zinc (ppm)	2053	631.5	29.00	493.7	69.18	18.02	91	1.99	037.42	0
037	Zinc (ppm)	0042	523.5	133.0	493.7	69.18	18.02	91	0.43	037.42	1
037	Zinc (ppm)	0560	601.3	202.8	493.7	69.18	18.02	91	1.56	037.52	1
037	Zinc (ppm)	2246	5,930	155.5	493.7	69.18	18.02	91	78.58	037.33	2
038	Molybdenum (ppm)	0918	0.9998	0.0447	1.548	0.3481	0.0604	18	-1.58	038.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
038	Molybdenum (ppm)	0910	1.050	0.0600	1.548	0.3481	0.0604	18	-1.43	038.52	0
038	Molybdenum (ppm)	0038	1.085	0.0100	1.548	0.3481	0.0604	18	-1.33	038.43	0
038	Molybdenum (ppm)	0171	1.093	0.0560	1.548	0.3481	0.0604	18	-1.31	038.41	0
038	Molybdenum (ppm)	0345	1.470	0.0200	1.548	0.3481	0.0604	18	-0.22	038.43	0
038	Molybdenum (ppm)	0226	1.480	0.0600	1.548	0.3481	0.0604	18	-0.20	038.41	0
038	Molybdenum (ppm)	0169	1.530	0.1400	1.548	0.3481	0.0604	18	-0.05	038.43	0
038	Molybdenum (ppm)	0353	1.534	0.0699	1.548	0.3481	0.0604	18	-0.04	038.43	0
038	Molybdenum (ppm)	0407	1.545	0.0982	1.548	0.3481	0.0604	18	-0.01	038.41	0
038	Molybdenum (ppm)	0939	1.615	0.0700	1.548	0.3481	0.0604	18	0.19	038.53	0
038	Molybdenum (ppm)	0083	1.650	0.1000	1.548	0.3481	0.0604	18	0.29	038.43	0
038	Molybdenum (ppm)	0693	1.708	0.0102	1.548	0.3481	0.0604	18	0.46	038.42	0
038	Molybdenum (ppm)	0297	1.710	0.1600	1.548	0.3481	0.0604	18	0.47	038.43	0
038	Molybdenum (ppm)	0407	1.770	0.0490	1.548	0.3481	0.0604	18	0.64	038.43	0
038	Molybdenum (ppm)	0278	1.800	0.0200	1.548	0.3481	0.0604	18	0.72	038.42	0
038	Molybdenum (ppm)	0553	1.835	0.0100	1.548	0.3481	0.0604	18	0.82	038.53	0
038	Molybdenum (ppm)	0164	1.915	0.0100	1.548	0.3481	0.0604	18	1.05	038.53	0
038	Molybdenum (ppm)	0510	2.050	0.1000	1.548	0.3481	0.0604	18	1.44	038.43	0
038	Molybdenum (ppm)	0265	1.920	0.7200	1.548	0.3481	0.0604	18	1.07	038.42	1
040	Barium (ppm)	0918	9.481	0.0049				2		040.53	0
040	Barium (ppm)	0560	10.41	0.6400				2		040.52	0
041	Vanadium (ppm)	0553	1.640	0.0400				1		041.53	0
042	Chloride (%)	0948	1.294	0.0240				2		042.00	0
042	Chloride (%)	0297	1.415	0.0100				2		042.99	0
101	Choline Chloride (ppm)	0619	522.0	12.00				2		101.02	0
101	Choline Chloride (ppm)	0227	1,230	0.0000				2		101.99	0
102	Niacin (ppm)	0939	13.45	0.5100				3		102.02	0
102	Niacin (ppm)	0160	20.15	1.100				3		102.99	0
102	Niacin (ppm)	0227	63.80	3.400				3		102.01	0
103	Pantothenic Acid (ppm)	0227	16.60	0.0000				3		103.01	0
103	Pantothenic Acid (ppm)	0939	18.61	1.040				3		103.02	0
103	Pantothenic Acid (ppm)	0160	57.70	9.000				3		103.99	0
104	Riboflavin (ppm)	0160	1.770	0.1200	4.564	2.185	0.5640	5		104.99	0
104	Riboflavin (ppm)	0010	3.450	0.7000	4.564	2.185	0.5640	5		104.03	0
104	Riboflavin (ppm)	0939	4.410	0.3000	4.564	2.185	0.5640	5		104.03	0
104	Riboflavin (ppm)	0227	5.655	0.6300	4.564	2.185	0.5640	5		104.00	0
104	Riboflavin (ppm)	0171	7.535	1.070	4.564	2.185	0.5640	5		104.00	0
105	Thiamine (ppm)	0010	9.300	2.000				3		105.00	0
105	Thiamine (ppm)	0227	12.45	0.1000				3		105.01	0
105	Thiamine (ppm)	0939	13.56	0.1300				3		105.00	0
106	Vitamin A (KU / kg)	0676	17.70	7.200	33.09	10.78	4.397	22	-1.43	106.02	0
106	Vitamin A (KU / kg)	0098	19.65	5.300	33.09	10.78	4.397	22	-1.25	106.01	0
106	Vitamin A (KU / kg)	0035	22.60	0.0000	33.09	10.78	4.397	22	-0.97	106.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
106	Vitamin A (KU / kg)	2192	23.81	0.8200	33.09	10.78	4.397	22	-0.86	106.02	0
106	Vitamin A (KU / kg)	0014	24.45	3.100	33.09	10.78	4.397	22	-0.80	106.02	0
106	Vitamin A (KU / kg)	2053	25.35	1.700	33.09	10.78	4.397	22	-0.72	106.02	0
106	Vitamin A (KU / kg)	0910	27.30	6.000	33.09	10.78	4.397	22	-0.54	106.02	0
106	Vitamin A (KU / kg)	0689	27.65	0.7000	33.09	10.78	4.397	22	-0.50	106.02	0
106	Vitamin A (KU / kg)	0610	29.00	7.000	33.09	10.78	4.397	22	-0.38	106.02	0
106	Vitamin A (KU / kg)	0038	29.79	2.129	33.09	10.78	4.397	22	-0.31	106.02	0
106	Vitamin A (KU / kg)	0010	31.85	9.500	33.09	10.78	4.397	22	-0.11	106.02	0
106	Vitamin A (KU / kg)	0563	34.15	0.0239	33.09	10.78	4.397	22	0.10	106.02	0
106	Vitamin A (KU / kg)	0905	34.77	14.71	33.09	10.78	4.397	22	0.16	106.02	0
106	Vitamin A (KU / kg)	0004	35.34	1.960	33.09	10.78	4.397	22	0.21	106.00	0
106	Vitamin A (KU / kg)	0017	35.83	7.110	33.09	10.78	4.397	22	0.25	106.02	0
106	Vitamin A (KU / kg)	0148	35.87	7.702	33.09	10.78	4.397	22	0.26	106.02	0
106	Vitamin A (KU / kg)	0169	39.50	1.000	33.09	10.78	4.397	22	0.60	106.02	0
106	Vitamin A (KU / kg)	0512	39.58	1.400	33.09	10.78	4.397	22	0.60	106.02	0
106	Vitamin A (KU / kg)	0171	46.10	3.400	33.09	10.78	4.397	22	1.21	106.00	0
106	Vitamin A (KU / kg)	2144	49.12	0.0320	33.09	10.78	4.397	22	1.49	106.99	0
106	Vitamin A (KU / kg)	0675	50.07	2.460	33.09	10.78	4.397	22	1.58	106.02	0
106	Vitamin A (KU / kg)	0941	52.90	13.49	33.09	10.78	4.397	22	1.84	106.02	0
107	Vitamin B12 (ppb)	0227	13.35	1.300				2		107.00	0
107	Vitamin B12 (ppb)	0160	40.00	10.00				2		107.99	0
108	Vitamin D3 (KU / kg)	0160	1.080	0.1800	4.307	1.806	0.2857	7	-1.79	108.99	0
108	Vitamin D3 (KU / kg)	0610	2.900	0.2000	4.307	1.806	0.2857	7	-0.78	108.02	0
108	Vitamin D3 (KU / kg)	0010	4.530	0.3800	4.307	1.806	0.2857	7	0.12	108.99	0
108	Vitamin D3 (KU / kg)	0227	4.740	0.0200	4.307	1.806	0.2857	7	0.24	108.02	0
108	Vitamin D3 (KU / kg)	0619	4.780	0.0400	4.307	1.806	0.2857	7	0.26	108.99	0
108	Vitamin D3 (KU / kg)	0675	5.100	0.1800	4.307	1.806	0.2857	7	0.44	108.02	0
108	Vitamin D3 (KU / kg)	0169	6.500	1.000	4.307	1.806	0.2857	7	1.21	108.01	0
109	Vitamin E (IU / kg)	0169	208.0	2.000	261.9	36.89	9.182	20	-1.46	109.02	0
109	Vitamin E (IU / kg)	0905	220.0	2.000	261.9	36.89	9.182	20	-1.14	109.02	0
109	Vitamin E (IU / kg)	2103	228.5	1.000	261.9	36.89	9.182	20	-0.90	109.02	0
109	Vitamin E (IU / kg)	0619	234.0	12.00	261.9	36.89	9.182	20	-0.76	109.02	0
109	Vitamin E (IU / kg)	0910	236.5	3.000	261.9	36.89	9.182	20	-0.69	109.02	0
109	Vitamin E (IU / kg)	0036	238.1	7.300	261.9	36.89	9.182	20	-0.65	109.02	0
109	Vitamin E (IU / kg)	0098	240.0	0.0000	261.9	36.89	9.182	20	-0.59	109.02	0
109	Vitamin E (IU / kg)	0610	245.5	29.00	261.9	36.89	9.182	20	-0.44	109.02	0
109	Vitamin E (IU / kg)	0171	247.5	5.000	261.9	36.89	9.182	20	-0.39	109.99	0
109	Vitamin E (IU / kg)	0941	250.7	7.390	261.9	36.89	9.182	20	-0.30	109.02	0
109	Vitamin E (IU / kg)	2192	266.3	17.11	261.9	36.89	9.182	20	0.12	109.02	0
109	Vitamin E (IU / kg)	2144	267.7	4.360	261.9	36.89	9.182	20	0.16	109.99	0
109	Vitamin E (IU / kg)	0563	267.8	1.129	261.9	36.89	9.182	20	0.16	109.02	0
109	Vitamin E (IU / kg)	0096	272.4	11.98	261.9	36.89	9.182	20	0.28	109.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
109	Vitamin E (IU / kg)	0148	281.1	7.130	261.9	36.89	9.182	20	0.52	109.02	0
109	Vitamin E (IU / kg)	0590	288.0	8.000	261.9	36.89	9.182	20	0.71	109.02	0
109	Vitamin E (IU / kg)	0676	300.2	6.100	261.9	36.89	9.182	20	1.04	109.02	0
109	Vitamin E (IU / kg)	0675	311.2	8.640	261.9	36.89	9.182	20	1.34	109.02	0
109	Vitamin E (IU / kg)	0038	372.7	49.70	261.9	36.89	9.182	20	3.00	109.02	0
109	Vitamin E (IU / kg)	0199	389.5	0.8000	261.9	36.89	9.182	20	3.46	109.02	0
109	Vitamin E (IU / kg)	0027	228.0	64.35	261.9	36.89	9.182	20	-0.92	109.02	1
111	Vitamin C, Phosphorylated (ppm)	0160	< 0.6					0		111.98	5
112	Pyridoxine (µg / g)	0160	0.8350	0.0700				2		112.99	0
112	Pyridoxine (µg / g)	0227	3.385	0.1100				2		112.01	0
113	Folic Acid (ppm)	0939	1.020	0.1200				3		113.02	0
113	Folic Acid (ppm)	0227	1.350	0.0600				3		113.99	0
113	Folic Acid (ppm)	0160	1.695	0.2900				3		113.99	0
114	Biotin (ppm)	0010	6.150	0.1000	10.63	4.892	0.5496	4		114.99	0
114	Biotin (ppm)	0227	8.205	0.1900	10.63	4.892	0.5496	4		114.01	0
114	Biotin (ppm)	0096	10.78	0.7083	10.63	4.892	0.5496	4		114.01	0
114	Biotin (ppm)	0160	17.40	1.200	10.63	4.892	0.5496	4		114.99	0
115	Non Protein N (NPN) (%)	0098	0.5500	0.0200				1		115.00	0
118	Peroxide value (meq/kg)	2081	4.845	0.9500				1		118.99	0
120	Alanine (%)	2246	0.5244	0.0093	0.6695	0.0294	0.0129	29	-4.93	120.05	0
120	Alanine (%)	0682	0.6200	0.0000	0.6695	0.0294	0.0129	29	-1.68	120.00	0
120	Alanine (%)	0918	0.6264	0.0012	0.6695	0.0294	0.0129	29	-1.46	120.00	0
120	Alanine (%)	0910	0.6400	0.0200	0.6695	0.0294	0.0129	29	-1.00	120.00	0
120	Alanine (%)	2196	0.6410	0.0000	0.6695	0.0294	0.0129	29	-0.97	120.05	0
120	Alanine (%)	0171	0.6450	0.0700	0.6695	0.0294	0.0129	29	-0.83	120.00	0
120	Alanine (%)	0968	0.6450	0.0100	0.6695	0.0294	0.0129	29	-0.83	120.00	0
120	Alanine (%)	0148	0.6500	0.0200	0.6695	0.0294	0.0129	29	-0.66	120.05	0
120	Alanine (%)	0226	0.6509	0.0392	0.6695	0.0294	0.0129	29	-0.63	120.00	0
120	Alanine (%)	0354	0.6525	0.0010	0.6695	0.0294	0.0129	29	-0.58	120.00	0
120	Alanine (%)	0652	0.6600	0.0000	0.6695	0.0294	0.0129	29	-0.32	120.00	0
120	Alanine (%)	0610	0.6650	0.0100	0.6695	0.0294	0.0129	29	-0.15	120.05	0
120	Alanine (%)	0870	0.6661	0.0077	0.6695	0.0294	0.0129	29	-0.12	120.00	0
120	Alanine (%)	0859	0.6690	0.0120	0.6695	0.0294	0.0129	29	-0.02	120.00	0
120	Alanine (%)	0098	0.6740	0.0040	0.6695	0.0294	0.0129	29	0.15	120.02	0
120	Alanine (%)	0872	0.6745	0.0190	0.6695	0.0294	0.0129	29	0.17	120.00	0
120	Alanine (%)	0868	0.6750	0.0420	0.6695	0.0294	0.0129	29	0.19	120.00	0
120	Alanine (%)	0889	0.6775	0.0050	0.6695	0.0294	0.0129	29	0.27	120.99	0
120	Alanine (%)	0227	0.6800	0.0000	0.6695	0.0294	0.0129	29	0.36	120.99	0
120	Alanine (%)	2059	0.6810	0.0080	0.6695	0.0294	0.0129	29	0.39	120.00	0
120	Alanine (%)	0571	0.6825	0.0010	0.6695	0.0294	0.0129	29	0.44	120.00	0
120	Alanine (%)	0684	0.6900	0.0040	0.6695	0.0294	0.0129	29	0.70	120.00	0
120	Alanine (%)	0644	0.6905	0.0010	0.6695	0.0294	0.0129	29	0.72	120.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
120	Alanine (%)	0878	0.6935	0.0030	0.6695	0.0294	0.0129	29	0.82	120.00	0
120	Alanine (%)	0242	0.7000	0.0000	0.6695	0.0294	0.0129	29	1.04	120.05	0
120	Alanine (%)	0353	0.7050	0.0100	0.6695	0.0294	0.0129	29	1.21	120.00	0
120	Alanine (%)	0941	0.7055	0.0550	0.6695	0.0294	0.0129	29	1.23	120.00	0
120	Alanine (%)	0675	0.7100	0.0200	0.6695	0.0294	0.0129	29	1.38	120.00	0
120	Alanine (%)	0619	0.7195	0.0030	0.6695	0.0294	0.0129	29	1.70	120.00	0
120	Alanine (%)	0407	0.5900	0.0980	0.6695	0.0294	0.0129	29	-2.70	120.05	1
121	Arginine (%)	2246	0.2914	0.0039	0.6658	0.0387	0.0180	29	-9.69	121.05	0
121	Arginine (%)	0407	0.5745	0.0890	0.6658	0.0387	0.0180	29	-2.36	121.05	0
121	Arginine (%)	0353	0.5800	0.0600	0.6658	0.0387	0.0180	29	-2.22	121.00	0
121	Arginine (%)	0918	0.5956	0.0237	0.6658	0.0387	0.0180	29	-1.82	121.00	0
121	Arginine (%)	0682	0.6210	0.0000	0.6658	0.0387	0.0180	29	-1.16	121.00	0
121	Arginine (%)	0968	0.6350	0.0100	0.6658	0.0387	0.0180	29	-0.80	121.00	0
121	Arginine (%)	0870	0.6398	0.0022	0.6658	0.0387	0.0180	29	-0.67	121.00	0
121	Arginine (%)	0227	0.6400	0.0000	0.6658	0.0387	0.0180	29	-0.67	121.99	0
121	Arginine (%)	0652	0.6500	0.0200	0.6658	0.0387	0.0180	29	-0.41	121.00	0
121	Arginine (%)	0098	0.6505	0.0010	0.6658	0.0387	0.0180	29	-0.40	121.02	0
121	Arginine (%)	0859	0.6524	0.0108	0.6658	0.0387	0.0180	29	-0.35	121.00	0
121	Arginine (%)	0878	0.6665	0.0130	0.6658	0.0387	0.0180	29	0.02	121.00	0
121	Arginine (%)	0354	0.6700	0.0020	0.6658	0.0387	0.0180	29	0.11	121.00	0
121	Arginine (%)	0226	0.6709	0.0486	0.6658	0.0387	0.0180	29	0.13	121.00	0
121	Arginine (%)	2196	0.6730	0.0000	0.6658	0.0387	0.0180	29	0.19	121.05	0
121	Arginine (%)	0684	0.6745	0.0250	0.6658	0.0387	0.0180	29	0.23	121.00	0
121	Arginine (%)	0171	0.6750	0.0100	0.6658	0.0387	0.0180	29	0.24	121.00	0
121	Arginine (%)	0872	0.6795	0.0090	0.6658	0.0387	0.0180	29	0.35	121.00	0
121	Arginine (%)	0675	0.6800	0.0200	0.6658	0.0387	0.0180	29	0.37	121.00	0
121	Arginine (%)	0242	0.6850	0.0100	0.6658	0.0387	0.0180	29	0.50	121.05	0
121	Arginine (%)	0610	0.6850	0.0100	0.6658	0.0387	0.0180	29	0.50	121.05	0
121	Arginine (%)	2059	0.6890	0.0080	0.6658	0.0387	0.0180	29	0.60	121.00	0
121	Arginine (%)	0910	0.6950	0.0100	0.6658	0.0387	0.0180	29	0.76	121.00	0
121	Arginine (%)	0571	0.6955	0.0130	0.6658	0.0387	0.0180	29	0.77	121.00	0
121	Arginine (%)	0868	0.6955	0.0050	0.6658	0.0387	0.0180	29	0.77	121.00	0
121	Arginine (%)	0644	0.6960	0.0080	0.6658	0.0387	0.0180	29	0.78	121.00	0
121	Arginine (%)	0889	0.7100	0.0100	0.6658	0.0387	0.0180	29	1.14	121.99	0
121	Arginine (%)	0619	0.7275	0.0110	0.6658	0.0387	0.0180	29	1.60	121.00	0
121	Arginine (%)	0148	0.7450	0.0900	0.6658	0.0387	0.0180	29	2.05	121.05	0
121	Arginine (%)	0941	3,871	7,741	0.6658	0.0387	0.0180	29	100117.00	121.00	2
122	Aspartic (%)	0682	0.8160	0.0000	0.9094	0.0329	0.0164	29	-2.84	122.00	0
122	Aspartic (%)	0918	0.8479	0.0046	0.9094	0.0329	0.0164	29	-1.87	122.00	0
122	Aspartic (%)	0859	0.8555	0.0228	0.9094	0.0329	0.0164	29	-1.64	122.00	0
122	Aspartic (%)	0226	0.8744	0.0420	0.9094	0.0329	0.0164	29	-1.06	122.00	0
122	Aspartic (%)	0148	0.8800	0.0400	0.9094	0.0329	0.0164	29	-0.89	122.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
122	Aspartic (%)	0870	0.8929	0.0160	0.9094	0.0329	0.0164	29	-0.50	122.00	0
122	Aspartic (%)	0407	0.8935	0.0390	0.9094	0.0329	0.0164	29	-0.48	122.05	0
122	Aspartic (%)	0968	0.8950	0.0100	0.9094	0.0329	0.0164	29	-0.44	122.00	0
122	Aspartic (%)	0878	0.8955	0.0010	0.9094	0.0329	0.0164	29	-0.42	122.00	0
122	Aspartic (%)	0872	0.8965	0.0170	0.9094	0.0329	0.0164	29	-0.39	122.00	0
122	Aspartic (%)	0652	0.9000	0.0200	0.9094	0.0329	0.0164	29	-0.29	122.00	0
122	Aspartic (%)	2059	0.9005	0.0010	0.9094	0.0329	0.0164	29	-0.27	122.00	0
122	Aspartic (%)	0644	0.9010	0.0000	0.9094	0.0329	0.0164	29	-0.26	122.00	0
122	Aspartic (%)	0098	0.9040	0.0000	0.9094	0.0329	0.0164	29	-0.17	122.02	0
122	Aspartic (%)	0227	0.9050	0.0100	0.9094	0.0329	0.0164	29	-0.13	122.99	0
122	Aspartic (%)	0684	0.9065	0.0090	0.9094	0.0329	0.0164	29	-0.09	122.00	0
122	Aspartic (%)	0610	0.9100	0.0000	0.9094	0.0329	0.0164	29	0.02	122.05	0
122	Aspartic (%)	0571	0.9105	0.0070	0.9094	0.0329	0.0164	29	0.03	122.00	0
122	Aspartic (%)	0941	0.9130	0.0580	0.9094	0.0329	0.0164	29	0.11	122.00	0
122	Aspartic (%)	0354	0.9160	0.0000	0.9094	0.0329	0.0164	29	0.20	122.00	0
122	Aspartic (%)	0910	0.9200	0.0000	0.9094	0.0329	0.0164	29	0.32	122.00	0
122	Aspartic (%)	0868	0.9230	0.0000	0.9094	0.0329	0.0164	29	0.41	122.00	0
122	Aspartic (%)	0353	0.9400	0.0400	0.9094	0.0329	0.0164	29	0.93	122.00	0
122	Aspartic (%)	2196	0.9400	0.0000	0.9094	0.0329	0.0164	29	0.93	122.05	0
122	Aspartic (%)	0675	0.9450	0.0300	0.9094	0.0329	0.0164	29	1.08	122.00	0
122	Aspartic (%)	0242	0.9550	0.0100	0.9094	0.0329	0.0164	29	1.38	122.05	0
122	Aspartic (%)	0171	0.9800	0.0600	0.9094	0.0329	0.0164	29	2.14	122.00	0
122	Aspartic (%)	0619	1.002	0.0170	0.9094	0.0329	0.0164	29	2.79	122.00	0
122	Aspartic (%)	0889	1.200	0.0200	0.9094	0.0329	0.0164	29	8.82	122.99	0
124	Cysteine/Cystine (%)	2246	0.0537	0.0087	0.2291	0.0265	0.0097	30	-6.61	124.05	0
124	Cysteine/Cystine (%)	0407	0.1385	0.0236	0.2291	0.0265	0.0097	30	-3.41	124.05	0
124	Cysteine/Cystine (%)	0889	0.1675	0.0150	0.2291	0.0265	0.0097	30	-2.32	124.99	0
124	Cysteine/Cystine (%)	0353	0.1900	0.0000	0.2291	0.0265	0.0097	30	-1.47	124.00	0
124	Cysteine/Cystine (%)	0870	0.2059	0.0089	0.2291	0.0265	0.0097	30	-0.87	124.00	0
124	Cysteine/Cystine (%)	0859	0.2071	0.0057	0.2291	0.0265	0.0097	30	-0.83	124.00	0
124	Cysteine/Cystine (%)	0227	0.2100	0.0000	0.2291	0.0265	0.0097	30	-0.72	124.00	0
124	Cysteine/Cystine (%)	0226	0.2129	0.0105	0.2291	0.0265	0.0097	30	-0.61	124.00	0
124	Cysteine/Cystine (%)	0652	0.2200	0.0000	0.2291	0.0265	0.0097	30	-0.34	124.00	0
124	Cysteine/Cystine (%)	0878	0.2220	0.0020	0.2291	0.0265	0.0097	30	-0.27	124.00	0
124	Cysteine/Cystine (%)	2059	0.2220	0.0040	0.2291	0.0265	0.0097	30	-0.27	124.00	0
124	Cysteine/Cystine (%)	0910	0.2250	0.0100	0.2291	0.0265	0.0097	30	-0.15	124.00	0
124	Cysteine/Cystine (%)	0918	0.2250	0.0100	0.2291	0.0265	0.0097	30	-0.15	124.00	0
124	Cysteine/Cystine (%)	0684	0.2260	0.0200	0.2291	0.0265	0.0097	30	-0.12	124.00	0
124	Cysteine/Cystine (%)	0872	0.2285	0.0190	0.2291	0.0265	0.0097	30	-0.02	124.00	0
124	Cysteine/Cystine (%)	0571	0.2295	0.0050	0.2291	0.0265	0.0097	30	0.02	124.00	0
124	Cysteine/Cystine (%)	0171	0.2300	0.0200	0.2291	0.0265	0.0097	30	0.04	124.00	0
124	Cysteine/Cystine (%)	0968	0.2350	0.0100	0.2291	0.0265	0.0097	30	0.22	124.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
124	Cysteine/Cystine (%)	0242	0.2350	0.0100	0.2291	0.0265	0.0097	30	0.22	124.05	0
124	Cysteine/Cystine (%)	0644	0.2375	0.0310	0.2291	0.0265	0.0097	30	0.32	124.00	0
124	Cysteine/Cystine (%)	0098	0.2395	0.0010	0.2291	0.0265	0.0097	30	0.39	124.02	0
124	Cysteine/Cystine (%)	2196	0.2410	0.0000	0.2291	0.0265	0.0097	30	0.45	124.05	0
124	Cysteine/Cystine (%)	0941	0.2440	0.0200	0.2291	0.0265	0.0097	30	0.56	124.00	0
124	Cysteine/Cystine (%)	0354	0.2465	0.0050	0.2291	0.0265	0.0097	30	0.66	124.00	0
124	Cysteine/Cystine (%)	0868	0.2500	0.0000	0.2291	0.0265	0.0097	30	0.79	124.00	0
124	Cysteine/Cystine (%)	0610	0.2550	0.0100	0.2291	0.0265	0.0097	30	0.98	124.05	0
124	Cysteine/Cystine (%)	0619	0.2600	0.0120	0.2291	0.0265	0.0097	30	1.17	124.00	0
124	Cysteine/Cystine (%)	0682	0.2810	0.0000	0.2291	0.0265	0.0097	30	1.96	124.00	0
124	Cysteine/Cystine (%)	0148	0.2900	0.0200	0.2291	0.0265	0.0097	30	2.30	124.05	0
124	Cysteine/Cystine (%)	0675	0.3050	0.0100	0.2291	0.0265	0.0097	30	2.86	124.00	0
125	Glutamic (%)	2246	0.9770	0.0117	2.111	0.1078	0.0475	29	-10.52	125.05	0
125	Glutamic (%)	0889	1.848	0.0250	2.111	0.1078	0.0475	29	-2.44	125.99	0
125	Glutamic (%)	0918	1.898	0.0706	2.111	0.1078	0.0475	29	-1.97	125.00	0
125	Glutamic (%)	0407	1.916	0.2790	2.111	0.1078	0.0475	29	-1.81	125.05	0
125	Glutamic (%)	0859	2.029	0.0474	2.111	0.1078	0.0475	29	-0.76	125.00	0
125	Glutamic (%)	0682	2.032	0.0000	2.111	0.1078	0.0475	29	-0.73	125.00	0
125	Glutamic (%)	2196	2.059	0.0000	2.111	0.1078	0.0475	29	-0.48	125.05	0
125	Glutamic (%)	0872	2.066	0.0760	2.111	0.1078	0.0475	29	-0.41	125.00	0
125	Glutamic (%)	0354	2.067	0.0090	2.111	0.1078	0.0475	29	-0.41	125.00	0
125	Glutamic (%)	0227	2.085	0.0100	2.111	0.1078	0.0475	29	-0.24	125.99	0
125	Glutamic (%)	0610	2.090	0.0200	2.111	0.1078	0.0475	29	-0.19	125.05	0
125	Glutamic (%)	0868	2.091	0.0130	2.111	0.1078	0.0475	29	-0.19	125.00	0
125	Glutamic (%)	0098	2.092	0.0090	2.111	0.1078	0.0475	29	-0.18	125.02	0
125	Glutamic (%)	2059	2.093	0.0100	2.111	0.1078	0.0475	29	-0.16	125.00	0
125	Glutamic (%)	0968	2.100	0.0400	2.111	0.1078	0.0475	29	-0.10	125.00	0
125	Glutamic (%)	0571	2.104	0.0570	2.111	0.1078	0.0475	29	-0.07	125.00	0
125	Glutamic (%)	0870	2.112	0.0127	2.111	0.1078	0.0475	29	0.01	125.00	0
125	Glutamic (%)	0171	2.125	0.1500	2.111	0.1078	0.0475	29	0.13	125.00	0
125	Glutamic (%)	0644	2.133	0.0250	2.111	0.1078	0.0475	29	0.20	125.00	0
125	Glutamic (%)	0684	2.143	0.0340	2.111	0.1078	0.0475	29	0.30	125.00	0
125	Glutamic (%)	0910	2.180	0.0400	2.111	0.1078	0.0475	29	0.64	125.00	0
125	Glutamic (%)	0941	2.183	0.1770	2.111	0.1078	0.0475	29	0.67	125.00	0
125	Glutamic (%)	0226	2.183	0.1297	2.111	0.1078	0.0475	29	0.67	125.00	0
125	Glutamic (%)	0242	2.190	0.0000	2.111	0.1078	0.0475	29	0.74	125.05	0
125	Glutamic (%)	0878	2.208	0.0120	2.111	0.1078	0.0475	29	0.90	125.00	0
125	Glutamic (%)	0148	2.235	0.0500	2.111	0.1078	0.0475	29	1.15	125.05	0
125	Glutamic (%)	0675	2.290	0.0600	2.111	0.1078	0.0475	29	1.66	125.00	0
125	Glutamic (%)	0353	2.440	0.0000	2.111	0.1078	0.0475	29	3.05	125.00	0
125	Glutamic (%)	0619	2.535	0.0100	2.111	0.1078	0.0475	29	3.94	125.00	0
125	Glutamic (%)	0652	1.065	2.050	2.111	0.1078	0.0475	29	-9.70	125.00	1

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
126	Glycine (%)	0889	0.3375	0.0050	0.5405	0.0304	0.0124	29	-6.67	126.99	0
126	Glycine (%)	0682	0.4900	0.0000	0.5405	0.0304	0.0124	29	-1.66	126.00	0
126	Glycine (%)	0918	0.5044	0.0083	0.5405	0.0304	0.0124	29	-1.19	126.00	0
126	Glycine (%)	0868	0.5065	0.0310	0.5405	0.0304	0.0124	29	-1.12	126.00	0
126	Glycine (%)	0870	0.5091	0.0054	0.5405	0.0304	0.0124	29	-1.03	126.00	0
126	Glycine (%)	0968	0.5150	0.0100	0.5405	0.0304	0.0124	29	-0.84	126.00	0
126	Glycine (%)	0859	0.5184	0.0107	0.5405	0.0304	0.0124	29	-0.73	126.00	0
126	Glycine (%)	0652	0.5200	0.0200	0.5405	0.0304	0.0124	29	-0.67	126.00	0
126	Glycine (%)	0354	0.5250	0.0040	0.5405	0.0304	0.0124	29	-0.51	126.00	0
126	Glycine (%)	0910	0.5250	0.0100	0.5405	0.0304	0.0124	29	-0.51	126.00	0
126	Glycine (%)	2196	0.5350	0.0000	0.5405	0.0304	0.0124	29	-0.18	126.05	0
126	Glycine (%)	0227	0.5350	0.0100	0.5405	0.0304	0.0124	29	-0.18	126.99	0
126	Glycine (%)	0878	0.5380	0.0000	0.5405	0.0304	0.0124	29	-0.08	126.00	0
126	Glycine (%)	0644	0.5400	0.0000	0.5405	0.0304	0.0124	29	-0.02	126.00	0
126	Glycine (%)	0571	0.5415	0.0090	0.5405	0.0304	0.0124	29	0.03	126.00	0
126	Glycine (%)	0872	0.5422	0.0003	0.5405	0.0304	0.0124	29	0.05	126.00	0
126	Glycine (%)	0098	0.5425	0.0030	0.5405	0.0304	0.0124	29	0.07	126.02	0
126	Glycine (%)	0675	0.5450	0.0100	0.5405	0.0304	0.0124	29	0.15	126.00	0
126	Glycine (%)	0610	0.5450	0.0100	0.5405	0.0304	0.0124	29	0.15	126.05	0
126	Glycine (%)	2059	0.5455	0.0070	0.5405	0.0304	0.0124	29	0.16	126.00	0
126	Glycine (%)	0226	0.5528	0.0415	0.5405	0.0304	0.0124	29	0.40	126.00	0
126	Glycine (%)	0148	0.5600	0.0600	0.5405	0.0304	0.0124	29	0.64	126.05	0
126	Glycine (%)	0941	0.5655	0.0410	0.5405	0.0304	0.0124	29	0.82	126.00	0
126	Glycine (%)	0242	0.5700	0.0000	0.5405	0.0304	0.0124	29	0.97	126.05	0
126	Glycine (%)	0171	0.5700	0.0200	0.5405	0.0304	0.0124	29	0.97	126.00	0
126	Glycine (%)	2246	0.5756	0.0073	0.5405	0.0304	0.0124	29	1.15	126.05	0
126	Glycine (%)	0407	0.5870	0.0160	0.5405	0.0304	0.0124	29	1.53	126.05	0
126	Glycine (%)	0353	0.6050	0.0100	0.5405	0.0304	0.0124	29	2.12	126.00	0
126	Glycine (%)	0619	0.6120	0.0100	0.5405	0.0304	0.0124	29	2.35	126.00	0
126	Glycine (%)	0684	0.5605	0.0710	0.5405	0.0304	0.0124	29	0.66	126.00	1
127	Histidine (%)	2246	0.1660	0.0193	0.2988	0.0214	0.0073	29	-6.21	127.05	0
127	Histidine (%)	2196	0.2630	0.0000	0.2988	0.0214	0.0073	29	-1.67	127.05	0
127	Histidine (%)	0870	0.2671	0.0183	0.2988	0.0214	0.0073	29	-1.48	127.00	0
127	Histidine (%)	0918	0.2718	0.0130	0.2988	0.0214	0.0073	29	-1.26	127.00	0
127	Histidine (%)	0226	0.2725	0.0213	0.2988	0.0214	0.0073	29	-1.23	127.00	0
127	Histidine (%)	0859	0.2791	0.0042	0.2988	0.0214	0.0073	29	-0.92	127.00	0
127	Histidine (%)	0610	0.2850	0.0100	0.2988	0.0214	0.0073	29	-0.64	127.05	0
127	Histidine (%)	0675	0.2900	0.0000	0.2988	0.0214	0.0073	29	-0.41	127.00	0
127	Histidine (%)	0684	0.2920	0.0080	0.2988	0.0214	0.0073	29	-0.32	127.00	0
127	Histidine (%)	0872	0.2945	0.0030	0.2988	0.0214	0.0073	29	-0.20	127.00	0
127	Histidine (%)	0968	0.2950	0.0100	0.2988	0.0214	0.0073	29	-0.18	127.00	0
127	Histidine (%)	0098	0.2955	0.0010	0.2988	0.0214	0.0073	29	-0.15	127.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
127	Histidine (%)	0354	0.2960	0.0020	0.2988	0.0214	0.0073	29	-0.13	127.00	0
127	Histidine (%)	0878	0.2980	0.0040	0.2988	0.0214	0.0073	29	-0.04	127.00	0
127	Histidine (%)	2059	0.2985	0.0010	0.2988	0.0214	0.0073	29	-0.01	127.00	0
127	Histidine (%)	0910	0.3000	0.0000	0.2988	0.0214	0.0073	29	0.06	127.00	0
127	Histidine (%)	0571	0.3005	0.0030	0.2988	0.0214	0.0073	29	0.08	127.00	0
127	Histidine (%)	0868	0.3035	0.0010	0.2988	0.0214	0.0073	29	0.22	127.00	0
127	Histidine (%)	0171	0.3050	0.0100	0.2988	0.0214	0.0073	29	0.29	127.00	0
127	Histidine (%)	0652	0.3050	0.0100	0.2988	0.0214	0.0073	29	0.29	127.00	0
127	Histidine (%)	0644	0.3055	0.0010	0.2988	0.0214	0.0073	29	0.31	127.00	0
127	Histidine (%)	0941	0.3175	0.0230	0.2988	0.0214	0.0073	29	0.87	127.00	0
127	Histidine (%)	0682	0.3180	0.0000	0.2988	0.0214	0.0073	29	0.90	127.00	0
127	Histidine (%)	0353	0.3200	0.0000	0.2988	0.0214	0.0073	29	0.99	127.00	0
127	Histidine (%)	0148	0.3200	0.0400	0.2988	0.0214	0.0073	29	0.99	127.05	0
127	Histidine (%)	0242	0.3200	0.0000	0.2988	0.0214	0.0073	29	0.99	127.05	0
127	Histidine (%)	0227	0.3200	0.0000	0.2988	0.0214	0.0073	29	0.99	127.99	0
127	Histidine (%)	0889	0.3375	0.0050	0.2988	0.0214	0.0073	29	1.81	127.99	0
127	Histidine (%)	0619	0.4110	0.0040	0.2988	0.0214	0.0073	29	5.24	127.00	0
127	Histidine (%)	0407	0.2645	0.0510	0.2988	0.0214	0.0073	29	-1.60	127.05	1
128	Isoleucine (%)	2246	0.2192	0.0039	0.3862	0.0317	0.0113	29	-5.28	128.05	0
128	Isoleucine (%)	0353	0.3000	0.0200	0.3862	0.0317	0.0113	29	-2.72	128.00	0
128	Isoleucine (%)	0682	0.3470	0.0000	0.3862	0.0317	0.0113	29	-1.24	128.00	0
128	Isoleucine (%)	0407	0.3495	0.0570	0.3862	0.0317	0.0113	29	-1.16	128.05	0
128	Isoleucine (%)	0918	0.3598	0.0037	0.3862	0.0317	0.0113	29	-0.83	128.00	0
128	Isoleucine (%)	0968	0.3600	0.0000	0.3862	0.0317	0.0113	29	-0.83	128.00	0
128	Isoleucine (%)	0619	0.3625	0.0270	0.3862	0.0317	0.0113	29	-0.75	128.00	0
128	Isoleucine (%)	0171	0.3650	0.0300	0.3862	0.0317	0.0113	29	-0.67	128.00	0
128	Isoleucine (%)	0868	0.3700	0.0060	0.3862	0.0317	0.0113	29	-0.51	128.00	0
128	Isoleucine (%)	0889	0.3700	0.0000	0.3862	0.0317	0.0113	29	-0.51	128.99	0
128	Isoleucine (%)	0684	0.3715	0.0030	0.3862	0.0317	0.0113	29	-0.46	128.00	0
128	Isoleucine (%)	0652	0.3750	0.0100	0.3862	0.0317	0.0113	29	-0.35	128.00	0
128	Isoleucine (%)	0870	0.3779	0.0111	0.3862	0.0317	0.0113	29	-0.26	128.00	0
128	Isoleucine (%)	0354	0.3785	0.0010	0.3862	0.0317	0.0113	29	-0.24	128.00	0
128	Isoleucine (%)	0878	0.3875	0.0010	0.3862	0.0317	0.0113	29	0.04	128.00	0
128	Isoleucine (%)	0098	0.3875	0.0070	0.3862	0.0317	0.0113	29	0.04	128.02	0
128	Isoleucine (%)	0872	0.3935	0.0010	0.3862	0.0317	0.0113	29	0.23	128.00	0
128	Isoleucine (%)	0227	0.3950	0.0100	0.3862	0.0317	0.0113	29	0.28	128.99	0
128	Isoleucine (%)	0910	0.4000	0.0000	0.3862	0.0317	0.0113	29	0.44	128.00	0
128	Isoleucine (%)	0859	0.4044	0.0073	0.3862	0.0317	0.0113	29	0.57	128.00	0
128	Isoleucine (%)	0610	0.4050	0.0100	0.3862	0.0317	0.0113	29	0.59	128.05	0
128	Isoleucine (%)	2059	0.4085	0.0010	0.3862	0.0317	0.0113	29	0.71	128.00	0
128	Isoleucine (%)	0571	0.4090	0.0000	0.3862	0.0317	0.0113	29	0.72	128.00	0
128	Isoleucine (%)	2196	0.4120	0.0000	0.3862	0.0317	0.0113	29	0.82	128.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
128	Isoleucine (%)	0941	0.4185	0.0350	0.3862	0.0317	0.0113	29	1.02	128.00	0
128	Isoleucine (%)	0644	0.4220	0.0020	0.3862	0.0317	0.0113	29	1.13	128.00	0
128	Isoleucine (%)	0242	0.4250	0.0300	0.3862	0.0317	0.0113	29	1.23	128.05	0
128	Isoleucine (%)	0148	0.4350	0.0300	0.3862	0.0317	0.0113	29	1.54	128.05	0
128	Isoleucine (%)	0675	0.4400	0.0200	0.3862	0.0317	0.0113	29	1.70	128.00	0
128	Isoleucine (%)	0226	0.3271	0.0936	0.3862	0.0317	0.0113	29	-1.87	128.00	1
129	Leucine (%)	2246	0.5714	0.0050	0.9724	0.0354	0.0212	30	-11.34	129.05	0
129	Leucine (%)	0889	0.7750	0.0100	0.9724	0.0354	0.0212	30	-5.58	129.99	0
129	Leucine (%)	0407	0.8005	0.0910	0.9724	0.0354	0.0212	30	-4.86	129.05	0
129	Leucine (%)	0682	0.8580	0.0000	0.9724	0.0354	0.0212	30	-3.24	129.00	0
129	Leucine (%)	0918	0.9443	0.0111	0.9724	0.0354	0.0212	30	-0.80	129.00	0
129	Leucine (%)	0652	0.9450	0.0100	0.9724	0.0354	0.0212	30	-0.78	129.00	0
129	Leucine (%)	0684	0.9495	0.0130	0.9724	0.0354	0.0212	30	-0.65	129.00	0
129	Leucine (%)	0968	0.9500	0.0000	0.9724	0.0354	0.0212	30	-0.63	129.00	0
129	Leucine (%)	2196	0.9510	0.0000	0.9724	0.0354	0.0212	30	-0.61	129.05	0
129	Leucine (%)	0354	0.9555	0.0010	0.9724	0.0354	0.0212	30	-0.48	129.00	0
129	Leucine (%)	0859	0.9653	0.0182	0.9724	0.0354	0.0212	30	-0.20	129.00	0
129	Leucine (%)	0878	0.9665	0.0090	0.9724	0.0354	0.0212	30	-0.17	129.00	0
129	Leucine (%)	0870	0.9676	0.0199	0.9724	0.0354	0.0212	30	-0.14	129.00	0
129	Leucine (%)	0171	0.9700	0.0200	0.9724	0.0354	0.0212	30	-0.07	129.00	0
129	Leucine (%)	0872	0.9735	0.0290	0.9724	0.0354	0.0212	30	0.03	129.00	0
129	Leucine (%)	0227	0.9750	0.0100	0.9724	0.0354	0.0212	30	0.07	129.99	0
129	Leucine (%)	0868	0.9765	0.0110	0.9724	0.0354	0.0212	30	0.11	129.00	0
129	Leucine (%)	0571	0.9770	0.0040	0.9724	0.0354	0.0212	30	0.13	129.00	0
129	Leucine (%)	0098	0.9775	0.0190	0.9724	0.0354	0.0212	30	0.14	129.02	0
129	Leucine (%)	0910	0.9800	0.0000	0.9724	0.0354	0.0212	30	0.21	129.00	0
129	Leucine (%)	0148	0.9950	0.0700	0.9724	0.0354	0.0212	30	0.64	129.05	0
129	Leucine (%)	0610	0.9950	0.0100	0.9724	0.0354	0.0212	30	0.64	129.05	0
129	Leucine (%)	0226	0.9956	0.0729	0.9724	0.0354	0.0212	30	0.65	129.00	0
129	Leucine (%)	2059	1.001	0.0110	0.9724	0.0354	0.0212	30	0.79	129.00	0
129	Leucine (%)	0941	1.004	0.0840	0.9724	0.0354	0.0212	30	0.89	129.00	0
129	Leucine (%)	0242	1.005	0.0300	0.9724	0.0354	0.0212	30	0.92	129.05	0
129	Leucine (%)	0644	1.006	0.0170	0.9724	0.0354	0.0212	30	0.93	129.00	0
129	Leucine (%)	0353	1.020	0.0000	0.9724	0.0354	0.0212	30	1.34	129.00	0
129	Leucine (%)	0619	1.075	0.0300	0.9724	0.0354	0.0212	30	2.90	129.00	0
129	Leucine (%)	0675	1.075	0.0300	0.9724	0.0354	0.0212	30	2.90	129.00	0
130	L-Lysine (%)	0407	0.0370	0.0020	0.5411	0.0362	0.0130	36	-13.91	130.05	0
130	L-Lysine (%)	2246	0.2277	0.0066	0.5411	0.0362	0.0130	36	-8.65	130.05	0
130	L-Lysine (%)	0889	0.4250	0.0300	0.5411	0.0362	0.0130	36	-3.20	130.99	0
130	L-Lysine (%)	2053	0.4695	0.0050	0.5411	0.0362	0.0130	36	-1.98	130.99	0
130	L-Lysine (%)	0918	0.4900	0.0170	0.5411	0.0362	0.0130	36	-1.41	130.00	0
130	L-Lysine (%)	0859	0.5105	0.0111	0.5411	0.0362	0.0130	36	-0.85	130.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
130	L-Lysine (%)	0870	0.5105	0.0086	0.5411	0.0362	0.0130	36	-0.85	130.00	0
130	L-Lysine (%)	0652	0.5150	0.0100	0.5411	0.0362	0.0130	36	-0.72	130.00	0
130	L-Lysine (%)	0027	0.5205	0.0010	0.5411	0.0362	0.0130	36	-0.57	130.05	0
130	L-Lysine (%)	0512	0.5210	0.0028	0.5411	0.0362	0.0130	36	-0.56	130.00	0
130	L-Lysine (%)	0148	0.5250	0.0100	0.5411	0.0362	0.0130	36	-0.44	130.05	0
130	L-Lysine (%)	0171	0.5300	0.0400	0.5411	0.0362	0.0130	36	-0.31	130.00	0
130	L-Lysine (%)	0878	0.5310	0.0040	0.5411	0.0362	0.0130	36	-0.28	130.00	0
130	L-Lysine (%)	0872	0.5335	0.0030	0.5411	0.0362	0.0130	36	-0.21	130.00	0
130	L-Lysine (%)	0910	0.5350	0.0100	0.5411	0.0362	0.0130	36	-0.17	130.00	0
130	L-Lysine (%)	0684	0.5385	0.0110	0.5411	0.0362	0.0130	36	-0.07	130.00	0
130	L-Lysine (%)	0868	0.5390	0.0180	0.5411	0.0362	0.0130	36	-0.06	130.00	0
130	L-Lysine (%)	0848	0.5400	0.0000	0.5411	0.0362	0.0130	36	-0.03	130.00	0
130	L-Lysine (%)	0968	0.5450	0.0100	0.5411	0.0362	0.0130	36	0.11	130.00	0
130	L-Lysine (%)	0610	0.5450	0.0100	0.5411	0.0362	0.0130	36	0.11	130.05	0
130	L-Lysine (%)	2196	0.5450	0.0000	0.5411	0.0362	0.0130	36	0.11	130.05	0
130	L-Lysine (%)	0939	0.5500	0.0200	0.5411	0.0362	0.0130	36	0.24	130.00	0
130	L-Lysine (%)	2059	0.5500	0.0040	0.5411	0.0362	0.0130	36	0.24	130.00	0
130	L-Lysine (%)	0644	0.5535	0.0090	0.5411	0.0362	0.0130	36	0.34	130.00	0
130	L-Lysine (%)	0354	0.5540	0.0020	0.5411	0.0362	0.0130	36	0.36	130.00	0
130	L-Lysine (%)	0353	0.5550	0.0300	0.5411	0.0362	0.0130	36	0.38	130.00	0
130	L-Lysine (%)	0619	0.5600	0.0020	0.5411	0.0362	0.0130	36	0.52	130.00	0
130	L-Lysine (%)	0941	0.5690	0.0500	0.5411	0.0362	0.0130	36	0.77	130.00	0
130	L-Lysine (%)	0014	0.5695	0.0510	0.5411	0.0362	0.0130	36	0.78	130.05	0
130	L-Lysine (%)	0675	0.5700	0.0200	0.5411	0.0362	0.0130	36	0.80	130.00	0
130	L-Lysine (%)	0571	0.5715	0.0070	0.5411	0.0362	0.0130	36	0.84	130.00	0
130	L-Lysine (%)	2144	0.5750	0.0100	0.5411	0.0362	0.0130	36	0.93	130.99	0
130	L-Lysine (%)	0242	0.5950	0.0100	0.5411	0.0362	0.0130	36	1.49	130.05	0
130	L-Lysine (%)	0682	0.6000	0.0000	0.5411	0.0362	0.0130	36	1.62	130.00	0
130	L-Lysine (%)	0098	0.6190	0.0140	0.5411	0.0362	0.0130	36	2.15	130.02	0
130	L-Lysine (%)	0227	0.6550	0.0300	0.5411	0.0362	0.0130	36	3.14	130.99	0
130	L-Lysine (%)	0226	0.5301	0.0735	0.5411	0.0362	0.0130	36	-0.31	130.00	1
131	Methionine (%)	2246	0.0772	0.0080	0.2028	0.0187	0.0089	29	-6.72	131.05	0
131	Methionine (%)	0148	0.1500	0.0400	0.2028	0.0187	0.0089	29	-2.83	131.05	0
131	Methionine (%)	0171	0.1600	0.0200	0.2028	0.0187	0.0089	29	-2.29	131.00	0
131	Methionine (%)	0889	0.1625	0.0150	0.2028	0.0187	0.0089	29	-2.16	131.99	0
131	Methionine (%)	0870	0.1846	0.0130	0.2028	0.0187	0.0089	29	-0.98	131.00	0
131	Methionine (%)	0407	0.1873	0.0321	0.2028	0.0187	0.0089	29	-0.83	131.05	0
131	Methionine (%)	0354	0.1885	0.0050	0.2028	0.0187	0.0089	29	-0.77	131.00	0
131	Methionine (%)	0859	0.1915	0.0038	0.2028	0.0187	0.0089	29	-0.61	131.00	0
131	Methionine (%)	0878	0.1975	0.0030	0.2028	0.0187	0.0089	29	-0.29	131.00	0
131	Methionine (%)	0227	0.2000	0.0000	0.2028	0.0187	0.0089	29	-0.15	131.00	0
131	Methionine (%)	0652	0.2000	0.0000	0.2028	0.0187	0.0089	29	-0.15	131.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
131	Methionine (%)	0918	0.2000	0.0000	0.2028	0.0187	0.0089	29	-0.15	131.00	0
131	Methionine (%)	0098	0.2015	0.0090	0.2028	0.0187	0.0089	29	-0.07	131.02	0
131	Methionine (%)	2196	0.2020	0.0000	0.2028	0.0187	0.0089	29	-0.04	131.05	0
131	Methionine (%)	0968	0.2025	0.0010	0.2028	0.0187	0.0089	29	-0.02	131.00	0
131	Methionine (%)	0242	0.2050	0.0100	0.2028	0.0187	0.0089	29	0.12	131.05	0
131	Methionine (%)	0684	0.2065	0.0270	0.2028	0.0187	0.0089	29	0.20	131.00	0
131	Methionine (%)	2059	0.2085	0.0010	0.2028	0.0187	0.0089	29	0.30	131.00	0
131	Methionine (%)	0353	0.2100	0.0000	0.2028	0.0187	0.0089	29	0.38	131.00	0
131	Methionine (%)	0941	0.2100	0.0100	0.2028	0.0187	0.0089	29	0.38	131.00	0
131	Methionine (%)	0868	0.2110	0.0040	0.2028	0.0187	0.0089	29	0.44	131.00	0
131	Methionine (%)	0872	0.2120	0.0100	0.2028	0.0187	0.0089	29	0.49	131.00	0
131	Methionine (%)	0571	0.2150	0.0020	0.2028	0.0187	0.0089	29	0.65	131.00	0
131	Methionine (%)	0682	0.2180	0.0000	0.2028	0.0187	0.0089	29	0.81	131.00	0
131	Methionine (%)	0226	0.2215	0.0037	0.2028	0.0187	0.0089	29	1.00	131.00	0
131	Methionine (%)	0910	0.2250	0.0100	0.2028	0.0187	0.0089	29	1.19	131.00	0
131	Methionine (%)	0619	0.2260	0.0200	0.2028	0.0187	0.0089	29	1.24	131.00	0
131	Methionine (%)	0644	0.2285	0.0010	0.2028	0.0187	0.0089	29	1.37	131.00	0
131	Methionine (%)	0675	0.2450	0.0100	0.2028	0.0187	0.0089	29	2.26	131.00	0
131	Methionine (%)	0610	10.61	20.78	0.2028	0.0187	0.0089	29	556.67	131.05	2
132	Phenylalanine (%)	2246	0.2431	0.0159	0.5162	0.0255	0.0146	30	-10.73	132.05	0
132	Phenylalanine (%)	0407	0.3060	0.0220	0.5162	0.0255	0.0146	30	-8.26	132.05	0
132	Phenylalanine (%)	0682	0.3950	0.0000	0.5162	0.0255	0.0146	30	-4.76	132.00	0
132	Phenylalanine (%)	0652	0.4750	0.0100	0.5162	0.0255	0.0146	30	-1.62	132.00	0
132	Phenylalanine (%)	0968	0.4850	0.0100	0.5162	0.0255	0.0146	30	-1.23	132.00	0
132	Phenylalanine (%)	2196	0.4990	0.0000	0.5162	0.0255	0.0146	30	-0.68	132.05	0
132	Phenylalanine (%)	0684	0.5020	0.0060	0.5162	0.0255	0.0146	30	-0.56	132.00	0
132	Phenylalanine (%)	0227	0.5050	0.0100	0.5162	0.0255	0.0146	30	-0.44	132.99	0
132	Phenylalanine (%)	0226	0.5053	0.0443	0.5162	0.0255	0.0146	30	-0.43	132.00	0
132	Phenylalanine (%)	0171	0.5100	0.0000	0.5162	0.0255	0.0146	30	-0.24	132.00	0
132	Phenylalanine (%)	0870	0.5113	0.0298	0.5162	0.0255	0.0146	30	-0.19	132.00	0
132	Phenylalanine (%)	0872	0.5120	0.0100	0.5162	0.0255	0.0146	30	-0.17	132.00	0
132	Phenylalanine (%)	2059	0.5125	0.0070	0.5162	0.0255	0.0146	30	-0.15	132.00	0
132	Phenylalanine (%)	0878	0.5130	0.0020	0.5162	0.0255	0.0146	30	-0.13	132.00	0
132	Phenylalanine (%)	0675	0.5150	0.0100	0.5162	0.0255	0.0146	30	-0.05	132.00	0
132	Phenylalanine (%)	0098	0.5150	0.0060	0.5162	0.0255	0.0146	30	-0.05	132.02	0
132	Phenylalanine (%)	0619	0.5160	0.0180	0.5162	0.0255	0.0146	30	-0.01	132.00	0
132	Phenylalanine (%)	0859	0.5199	0.0101	0.5162	0.0255	0.0146	30	0.14	132.00	0
132	Phenylalanine (%)	0571	0.5220	0.0120	0.5162	0.0255	0.0146	30	0.23	132.00	0
132	Phenylalanine (%)	0910	0.5250	0.0100	0.5162	0.0255	0.0146	30	0.34	132.00	0
132	Phenylalanine (%)	0644	0.5275	0.0070	0.5162	0.0255	0.0146	30	0.44	132.00	0
132	Phenylalanine (%)	0354	0.5300	0.0020	0.5162	0.0255	0.0146	30	0.54	132.00	0
132	Phenylalanine (%)	0242	0.5350	0.0100	0.5162	0.0255	0.0146	30	0.74	132.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
132	Phenylalanine (%)	0610	0.5350	0.0100	0.5162	0.0255	0.0146	30	0.74	132.05	0
132	Phenylalanine (%)	0868	0.5360	0.0060	0.5162	0.0255	0.0146	30	0.78	132.00	0
132	Phenylalanine (%)	0889	0.5400	0.0100	0.5162	0.0255	0.0146	30	0.93	132.99	0
132	Phenylalanine (%)	0941	0.5445	0.0490	0.5162	0.0255	0.0146	30	1.11	132.00	0
132	Phenylalanine (%)	0148	0.5500	0.0600	0.5162	0.0255	0.0146	30	1.33	132.05	0
132	Phenylalanine (%)	0918	0.6029	0.0110	0.5162	0.0255	0.0146	30	3.40	132.00	0
132	Phenylalanine (%)	0353	0.8500	0.0400	0.5162	0.0255	0.0146	30	13.11	132.00	0
133	Proline (%)	0859	0.5078	0.0084	0.8154	0.0557	0.0206	28	-5.53	133.00	0
133	Proline (%)	2246	0.5207	0.0068	0.8154	0.0557	0.0206	28	-5.29	133.05	0
133	Proline (%)	0870	0.7469	0.0330	0.8154	0.0557	0.0206	28	-1.23	133.00	0
133	Proline (%)	0910	0.7500	0.0800	0.8154	0.0557	0.0206	28	-1.17	133.00	0
133	Proline (%)	0889	0.7550	0.0000	0.8154	0.0557	0.0206	28	-1.08	133.99	0
133	Proline (%)	0354	0.7655	0.0010	0.8154	0.0557	0.0206	28	-0.90	133.00	0
133	Proline (%)	0868	0.7795	0.0470	0.8154	0.0557	0.0206	28	-0.64	133.00	0
133	Proline (%)	0227	0.7800	0.0400	0.8154	0.0557	0.0206	28	-0.64	133.99	0
133	Proline (%)	0872	0.7970	0.0280	0.8154	0.0557	0.0206	28	-0.33	133.00	0
133	Proline (%)	0652	0.8000	0.0200	0.8154	0.0557	0.0206	28	-0.28	133.00	0
133	Proline (%)	0675	0.8050	0.0100	0.8154	0.0557	0.0206	28	-0.19	133.00	0
133	Proline (%)	0684	0.8060	0.0120	0.8154	0.0557	0.0206	28	-0.17	133.00	0
133	Proline (%)	0644	0.8095	0.0070	0.8154	0.0557	0.0206	28	-0.11	133.00	0
133	Proline (%)	2196	0.8130	0.0000	0.8154	0.0557	0.0206	28	-0.04	133.05	0
133	Proline (%)	0968	0.8250	0.0100	0.8154	0.0557	0.0206	28	0.17	133.00	0
133	Proline (%)	0571	0.8270	0.0260	0.8154	0.0557	0.0206	28	0.21	133.00	0
133	Proline (%)	0610	0.8300	0.0200	0.8154	0.0557	0.0206	28	0.26	133.05	0
133	Proline (%)	0878	0.8305	0.0030	0.8154	0.0557	0.0206	28	0.27	133.00	0
133	Proline (%)	0619	0.8360	0.0360	0.8154	0.0557	0.0206	28	0.37	133.00	0
133	Proline (%)	2059	0.8380	0.0000	0.8154	0.0557	0.0206	28	0.41	133.00	0
133	Proline (%)	0941	0.8445	0.0550	0.8154	0.0557	0.0206	28	0.52	133.00	0
133	Proline (%)	0171	0.8450	0.0500	0.8154	0.0557	0.0206	28	0.53	133.00	0
133	Proline (%)	0148	0.8450	0.0500	0.8154	0.0557	0.0206	28	0.53	133.05	0
133	Proline (%)	0242	0.8450	0.0100	0.8154	0.0557	0.0206	28	0.53	133.05	0
133	Proline (%)	0682	0.8970	0.0000	0.8154	0.0557	0.0206	28	1.47	133.00	0
133	Proline (%)	0353	0.9200	0.0000	0.8154	0.0557	0.0206	28	1.88	133.00	0
133	Proline (%)	0226	0.9258	0.0126	0.8154	0.0557	0.0206	28	1.98	133.00	0
133	Proline (%)	0918	1.008	0.0109	0.8154	0.0557	0.0206	28	3.46	133.00	0
133	Proline (%)	0407	0.7100	0.1320	0.8154	0.0557	0.0206	28	-1.89	133.05	1
134	Serine (%)	0407	0.3635	0.0130	0.5365	0.0365	0.0126	30	-4.73	134.05	0
134	Serine (%)	2246	0.4209	0.0057	0.5365	0.0365	0.0126	30	-3.17	134.05	0
134	Serine (%)	0098	0.4655	0.0050	0.5365	0.0365	0.0126	30	-1.94	134.02	0
134	Serine (%)	0918	0.4977	0.0001	0.5365	0.0365	0.0126	30	-1.06	134.00	0
134	Serine (%)	2059	0.5015	0.0070	0.5365	0.0365	0.0126	30	-0.96	134.00	0
134	Serine (%)	0644	0.5065	0.0090	0.5365	0.0365	0.0126	30	-0.82	134.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
134	Serine (%)	0610	0.5100	0.0000	0.5365	0.0365	0.0126	30	-0.73	134.05	0
134	Serine (%)	0652	0.5150	0.0300	0.5365	0.0365	0.0126	30	-0.59	134.00	0
134	Serine (%)	0859	0.5177	0.0116	0.5365	0.0365	0.0126	30	-0.51	134.00	0
134	Serine (%)	0872	0.5195	0.0210	0.5365	0.0365	0.0126	30	-0.47	134.00	0
134	Serine (%)	0878	0.5220	0.0020	0.5365	0.0365	0.0126	30	-0.40	134.00	0
134	Serine (%)	0227	0.5250	0.0100	0.5365	0.0365	0.0126	30	-0.31	134.99	0
134	Serine (%)	0571	0.5260	0.0040	0.5365	0.0365	0.0126	30	-0.29	134.00	0
134	Serine (%)	0968	0.5300	0.0000	0.5365	0.0365	0.0126	30	-0.18	134.00	0
134	Serine (%)	2196	0.5330	0.0000	0.5365	0.0365	0.0126	30	-0.10	134.05	0
134	Serine (%)	0941	0.5375	0.0290	0.5365	0.0365	0.0126	30	0.03	134.00	0
134	Serine (%)	0868	0.5425	0.0170	0.5365	0.0365	0.0126	30	0.16	134.00	0
134	Serine (%)	0682	0.5430	0.0000	0.5365	0.0365	0.0126	30	0.18	134.00	0
134	Serine (%)	0148	0.5450	0.0500	0.5365	0.0365	0.0126	30	0.23	134.05	0
134	Serine (%)	0242	0.5450	0.0100	0.5365	0.0365	0.0126	30	0.23	134.05	0
134	Serine (%)	0870	0.5612	0.0221	0.5365	0.0365	0.0126	30	0.67	134.00	0
134	Serine (%)	0171	0.5650	0.0300	0.5365	0.0365	0.0126	30	0.78	134.00	0
134	Serine (%)	0675	0.5650	0.0100	0.5365	0.0365	0.0126	30	0.78	134.00	0
134	Serine (%)	0226	0.5654	0.0196	0.5365	0.0365	0.0126	30	0.79	134.00	0
134	Serine (%)	0354	0.5655	0.0010	0.5365	0.0365	0.0126	30	0.79	134.00	0
134	Serine (%)	0684	0.5685	0.0210	0.5365	0.0365	0.0126	30	0.88	134.00	0
134	Serine (%)	0910	0.5700	0.0200	0.5365	0.0365	0.0126	30	0.92	134.00	0
134	Serine (%)	0353	0.5900	0.0000	0.5365	0.0365	0.0126	30	1.46	134.00	0
134	Serine (%)	0619	0.6195	0.0210	0.5365	0.0365	0.0126	30	2.27	134.00	0
134	Serine (%)	0889	0.7200	0.0100	0.5365	0.0365	0.0126	30	5.02	134.99	0
135	Threonine (%)	2246	0.2851	0.0002	0.4082	0.0224	0.0109	29	-5.49	135.05	0
135	Threonine (%)	0407	0.2905	0.0250	0.4082	0.0224	0.0109	29	-5.25	135.05	0
135	Threonine (%)	0682	0.3190	0.0000	0.4082	0.0224	0.0109	29	-3.98	135.00	0
135	Threonine (%)	0652	0.3850	0.0100	0.4082	0.0224	0.0109	29	-1.04	135.00	0
135	Threonine (%)	0918	0.3860	0.0026	0.4082	0.0224	0.0109	29	-0.99	135.00	0
135	Threonine (%)	0226	0.3914	0.0323	0.4082	0.0224	0.0109	29	-0.75	135.00	0
135	Threonine (%)	0859	0.3939	0.0055	0.4082	0.0224	0.0109	29	-0.64	135.00	0
135	Threonine (%)	2059	0.3955	0.0050	0.4082	0.0224	0.0109	29	-0.57	135.00	0
135	Threonine (%)	0610	0.4000	0.0000	0.4082	0.0224	0.0109	29	-0.37	135.05	0
135	Threonine (%)	0872	0.4010	0.0060	0.4082	0.0224	0.0109	29	-0.32	135.00	0
135	Threonine (%)	0098	0.4010	0.0020	0.4082	0.0224	0.0109	29	-0.32	135.02	0
135	Threonine (%)	0878	0.4030	0.0020	0.4082	0.0224	0.0109	29	-0.23	135.00	0
135	Threonine (%)	0571	0.4040	0.0020	0.4082	0.0224	0.0109	29	-0.19	135.00	0
135	Threonine (%)	0870	0.4066	0.0025	0.4082	0.0224	0.0109	29	-0.07	135.00	0
135	Threonine (%)	0644	0.4075	0.0070	0.4082	0.0224	0.0109	29	-0.03	135.00	0
135	Threonine (%)	0868	0.4075	0.0150	0.4082	0.0224	0.0109	29	-0.03	135.00	0
135	Threonine (%)	0227	0.4100	0.0000	0.4082	0.0224	0.0109	29	0.08	135.99	0
135	Threonine (%)	0968	0.4100	0.0200	0.4082	0.0224	0.0109	29	0.08	135.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
135	Threonine (%)	0354	0.4115	0.0010	0.4082	0.0224	0.0109	29	0.15	135.00	0
135	Threonine (%)	0171	0.4150	0.0300	0.4082	0.0224	0.0109	29	0.30	135.00	0
135	Threonine (%)	0941	0.4165	0.0270	0.4082	0.0224	0.0109	29	0.37	135.00	0
135	Threonine (%)	2196	0.4220	0.0000	0.4082	0.0224	0.0109	29	0.62	135.05	0
135	Threonine (%)	0910	0.4300	0.0200	0.4082	0.0224	0.0109	29	0.97	135.00	0
135	Threonine (%)	0148	0.4300	0.0400	0.4082	0.0224	0.0109	29	0.97	135.05	0
135	Threonine (%)	0242	0.4300	0.0200	0.4082	0.0224	0.0109	29	0.97	135.05	0
135	Threonine (%)	0675	0.4350	0.0100	0.4082	0.0224	0.0109	29	1.20	135.00	0
135	Threonine (%)	0684	0.4385	0.0170	0.4082	0.0224	0.0109	29	1.35	135.00	0
135	Threonine (%)	0619	0.4545	0.0090	0.4082	0.0224	0.0109	29	2.07	135.00	0
135	Threonine (%)	0889	0.5075	0.0050	0.4082	0.0224	0.0109	29	4.43	135.99	0
135	Threonine (%)	0353	0.4450	0.0700	0.4082	0.0224	0.0109	29	1.64	135.00	1
136	Tryptophan (%)	0171	0.0700	0.0200	0.1400	0.0188	0.0047	18	-3.72	136.00	0
136	Tryptophan (%)	0682	0.1180	0.0000	0.1400	0.0188	0.0047	18	-1.17	136.00	0
136	Tryptophan (%)	0910	0.1200	0.0000	0.1400	0.0188	0.0047	18	-1.06	136.00	0
136	Tryptophan (%)	0242	0.1250	0.0100	0.1400	0.0188	0.0047	18	-0.79	136.05	0
136	Tryptophan (%)	0859	0.1302	0.0028	0.1400	0.0188	0.0047	18	-0.52	136.03	0
136	Tryptophan (%)	0098	0.1325	0.0050	0.1400	0.0188	0.0047	18	-0.40	136.02	0
136	Tryptophan (%)	0878	0.1340	0.0020	0.1400	0.0188	0.0047	18	-0.32	136.01	0
136	Tryptophan (%)	0872	0.1380	0.0040	0.1400	0.0188	0.0047	18	-0.10	136.03	0
136	Tryptophan (%)	2059	0.1385	0.0030	0.1400	0.0188	0.0047	18	-0.08	136.03	0
136	Tryptophan (%)	0868	0.1395	0.0030	0.1400	0.0188	0.0047	18	-0.02	136.01	0
136	Tryptophan (%)	0571	0.1405	0.0030	0.1400	0.0188	0.0047	18	0.03	136.03	0
136	Tryptophan (%)	0644	0.1435	0.0010	0.1400	0.0188	0.0047	18	0.19	136.01	0
136	Tryptophan (%)	0870	0.1463	0.0063	0.1400	0.0188	0.0047	18	0.33	136.00	0
136	Tryptophan (%)	0684	0.1480	0.0140	0.1400	0.0188	0.0047	18	0.43	136.00	0
136	Tryptophan (%)	0619	0.1490	0.0060	0.1400	0.0188	0.0047	18	0.48	136.00	0
136	Tryptophan (%)	0227	0.1700	0.0000	0.1400	0.0188	0.0047	18	1.60	136.00	0
136	Tryptophan (%)	0918	0.1700	0.0000	0.1400	0.0188	0.0047	18	1.60	136.00	0
136	Tryptophan (%)	0889	0.2725	0.0050	0.1400	0.0188	0.0047	18	7.04	136.99	0
136	Tryptophan (%)	0353	0.1900	0.0800	0.1400	0.0188	0.0047	18	2.66	136.03	1
137	Tyrosine (%)	2246	0.1062	0.0111	0.3405	0.0518	0.0106	22	-4.52	137.05	0
137	Tyrosine (%)	0910	0.2600	0.0200	0.3405	0.0518	0.0106	22	-1.55	137.00	0
137	Tyrosine (%)	0889	0.2725	0.0050	0.3405	0.0518	0.0106	22	-1.31	137.99	0
137	Tyrosine (%)	0098	0.2850	0.0020	0.3405	0.0518	0.0106	22	-1.07	137.02	0
137	Tyrosine (%)	0227	0.3000	0.0000	0.3405	0.0518	0.0106	22	-0.78	137.99	0
137	Tyrosine (%)	0870	0.3196	0.0135	0.3405	0.0518	0.0106	22	-0.40	137.00	0
137	Tyrosine (%)	0171	0.3200	0.0200	0.3405	0.0518	0.0106	22	-0.40	137.00	0
137	Tyrosine (%)	0354	0.3225	0.0010	0.3405	0.0518	0.0106	22	-0.35	137.00	0
137	Tyrosine (%)	0226	0.3284	0.0045	0.3405	0.0518	0.0106	22	-0.23	137.00	0
137	Tyrosine (%)	0872	0.3300	0.0040	0.3405	0.0518	0.0106	22	-0.20	137.00	0
137	Tyrosine (%)	0968	0.3400	0.0000	0.3405	0.0518	0.0106	22	-0.01	137.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
137	Tyrosine (%)	0675	0.3450	0.0100	0.3405	0.0518	0.0106	22	0.09	137.00	0
137	Tyrosine (%)	0684	0.3475	0.0010	0.3405	0.0518	0.0106	22	0.14	137.00	0
137	Tyrosine (%)	2196	0.3550	0.0000	0.3405	0.0518	0.0106	22	0.28	137.05	0
137	Tyrosine (%)	0242	0.3650	0.0100	0.3405	0.0518	0.0106	22	0.47	137.05	0
137	Tyrosine (%)	0941	0.3690	0.0400	0.3405	0.0518	0.0106	22	0.55	137.00	0
137	Tyrosine (%)	0644	0.3710	0.0180	0.3405	0.0518	0.0106	22	0.59	137.00	0
137	Tyrosine (%)	0918	0.3787	0.0199	0.3405	0.0518	0.0106	22	0.74	137.00	0
137	Tyrosine (%)	2059	0.3900	0.0040	0.3405	0.0518	0.0106	22	0.96	137.00	0
137	Tyrosine (%)	0610	0.3900	0.0000	0.3405	0.0518	0.0106	22	0.96	137.05	0
137	Tyrosine (%)	0148	0.4550	0.0500	0.3405	0.0518	0.0106	22	2.21	137.05	0
137	Tyrosine (%)	0682	0.4940	0.0000	0.3405	0.0518	0.0106	22	2.96	137.00	0
137	Tyrosine (%)	0353	0.8550	0.0700	0.3405	0.0518	0.0106	22	9.93	137.00	1
138	Valine (%)	0226	0.3271	0.0936	0.5423	0.0436	0.0181	29	-4.94	138.00	0
138	Valine (%)	2246	0.3334	0.0045	0.5423	0.0436	0.0181	29	-4.79	138.05	0
138	Valine (%)	0870	0.4678	0.0330	0.5423	0.0436	0.0181	29	-1.71	138.00	0
138	Valine (%)	0171	0.4950	0.0100	0.5423	0.0436	0.0181	29	-1.08	138.00	0
138	Valine (%)	0918	0.5005	0.0034	0.5423	0.0436	0.0181	29	-0.96	138.00	0
138	Valine (%)	0619	0.5090	0.0380	0.5423	0.0436	0.0181	29	-0.76	138.00	0
138	Valine (%)	2196	0.5140	0.0000	0.5423	0.0436	0.0181	29	-0.65	138.05	0
138	Valine (%)	0652	0.5150	0.0100	0.5423	0.0436	0.0181	29	-0.63	138.00	0
138	Valine (%)	0353	0.5200	0.0400	0.5423	0.0436	0.0181	29	-0.51	138.00	0
138	Valine (%)	0682	0.5220	0.0000	0.5423	0.0436	0.0181	29	-0.47	138.00	0
138	Valine (%)	0684	0.5295	0.0470	0.5423	0.0436	0.0181	29	-0.29	138.00	0
138	Valine (%)	0868	0.5310	0.0080	0.5423	0.0436	0.0181	29	-0.26	138.00	0
138	Valine (%)	0941	0.5320	0.0560	0.5423	0.0436	0.0181	29	-0.24	138.00	0
138	Valine (%)	0354	0.5335	0.0030	0.5423	0.0436	0.0181	29	-0.20	138.00	0
138	Valine (%)	0872	0.5435	0.0010	0.5423	0.0436	0.0181	29	0.03	138.00	0
138	Valine (%)	0878	0.5440	0.0060	0.5423	0.0436	0.0181	29	0.04	138.00	0
138	Valine (%)	0910	0.5500	0.0200	0.5423	0.0436	0.0181	29	0.18	138.00	0
138	Valine (%)	0859	0.5536	0.0110	0.5423	0.0436	0.0181	29	0.26	138.00	0
138	Valine (%)	0968	0.5550	0.0100	0.5423	0.0436	0.0181	29	0.29	138.00	0
138	Valine (%)	0098	0.5630	0.0140	0.5423	0.0436	0.0181	29	0.48	138.02	0
138	Valine (%)	0571	0.5655	0.0070	0.5423	0.0436	0.0181	29	0.53	138.00	0
138	Valine (%)	2059	0.5685	0.0030	0.5423	0.0436	0.0181	29	0.60	138.00	0
138	Valine (%)	0644	0.5785	0.0070	0.5423	0.0436	0.0181	29	0.83	138.00	0
138	Valine (%)	0227	0.5850	0.0100	0.5423	0.0436	0.0181	29	0.98	138.99	0
138	Valine (%)	0675	0.5950	0.0100	0.5423	0.0436	0.0181	29	1.21	138.00	0
138	Valine (%)	0148	0.5950	0.0300	0.5423	0.0436	0.0181	29	1.21	138.05	0
138	Valine (%)	0242	0.5950	0.0300	0.5423	0.0436	0.0181	29	1.21	138.05	0
138	Valine (%)	0610	0.5950	0.0100	0.5423	0.0436	0.0181	29	1.21	138.05	0
138	Valine (%)	0889	0.6350	0.0100	0.5423	0.0436	0.0181	29	2.13	138.99	0
138	Valine (%)	0407	0.4890	0.1280	0.5423	0.0436	0.0181	29	-1.22	138.05	1

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
139	Taurine (%)	0407	0.0011	0.0019				3		139.05	0
139	Taurine (%)	0682	0.0600	0.0000				3		139.00	0
139	Taurine (%)	0941	0.1340	0.0080				3		139.00	0
139	Taurine (%)	0098	< 0.01					3		139.02	5
160	Fructose (%)	0227	0.4100	0.0200				3		160.99	0
160	Fructose (%)	0297	0.4340	0.0260				3		160.10	0
160	Fructose (%)	0619	0.4525	0.0110				3		160.99	0
160	Fructose (%)	0889	0.0000	0.0000				3		160.99	4
161	Galactose (%)	0297	0.0000	0.0000				0		161.10	4
162	Glucose (%)	0619	0.2520	0.0020				3		162.99	0
162	Glucose (%)	0227	0.2950	0.0100				3		162.99	0
162	Glucose (%)	0297	0.3005	0.0150				3		162.10	0
162	Glucose (%)	0889	0.0000	0.0000				3		162.99	4
163	Lactose (%)	0297	0.0855	0.0150				1		163.10	0
163	Lactose (%)	0619	0.0000	0.0000				1		163.99	4
163	Lactose (%)	0227	< 0.15					1		163.99	5
164	Maltose (%)	0297	0.2435	0.0370				1		164.10	0
164	Maltose (%)	0227	< 0.15					1		164.99	5
165	Sucrose (%)	0297	2.195	0.1010				3		165.10	0
165	Sucrose (%)	0227	2.360	0.0400				3		165.99	0
165	Sucrose (%)	0619	2.480	0.0600				3		165.99	0
165	Sucrose (%)	0889	0.0000	0.0000				3		165.99	4
166	Raffinose (%)	0297	0.2185	0.0230				3		166.10	0
166	Raffinose (%)	0227	0.2500	0.0000				3		166.99	0
166	Raffinose (%)	0619	0.3395	0.0090				3		166.99	0
167	Stachyose (%)	0297	0.0615	0.0130				2		167.10	0
167	Stachyose (%)	0619	0.1890	0.0060				2		167.99	0
167	Stachyose (%)	0227	< 0.05					2		167.99	5
314	Semduramicin Sodium (ppm)	0218	< 0.1					0		314.04	5
317	Maduramicin (ppm)	0218	< 0.1					0		317.05	5
351	Chlortetracycline (ppm)	0939	1.207	0.0770	2.393	0.9458	0.1709	5		351.05	0
351	Chlortetracycline (ppm)	0218	1.672	0.1405	2.393	0.9458	0.1709	5		351.00	0
351	Chlortetracycline (ppm)	0148	2.750	0.3070	2.393	0.9458	0.1709	5		351.05	0
351	Chlortetracycline (ppm)	0619	2.765	0.0300	2.393	0.9458	0.1709	5		351.05	0
351	Chlortetracycline (ppm)	0160	3.570	0.3000	2.393	0.9458	0.1709	5		351.05	0
355	Erythromycin (ppm)	0218	< 0.1					0		355.03	5
361	Lasalocid Sodium (ppm)	0218	< 0.1					0		361.05	5
363	Lincomycin (ppm)	0218	< 0.1					0		363.03	5
365	Monensin (ppm)	0009	1.707	0.2610	2.115	0.2210	0.1917	16	-1.85	365.05	0
365	Monensin (ppm)	0875	1.785	0.2100	2.115	0.2210	0.1917	16	-1.49	365.05	0
365	Monensin (ppm)	0905	1.950	0.1000	2.115	0.2210	0.1917	16	-0.75	365.02	0
365	Monensin (ppm)	0218	2.000	0.0400	2.115	0.2210	0.1917	16	-0.52	365.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
365	Monensin (ppm)	2053	2.010	0.1800	2.115	0.2210	0.1917	16	-0.48	365.05	0
365	Monensin (ppm)	0668	2.065	0.0500	2.115	0.2210	0.1917	16	-0.23	365.99	0
365	Monensin (ppm)	2192	2.085	0.0700	2.115	0.2210	0.1917	16	-0.14	365.03	0
365	Monensin (ppm)	0610	2.095	0.0100	2.115	0.2210	0.1917	16	-0.09	365.05	0
365	Monensin (ppm)	0668	2.100	0.0000	2.115	0.2210	0.1917	16	-0.07	365.02	0
365	Monensin (ppm)	0964	2.115	0.4156	2.115	0.2210	0.1917	16	0.00	365.03	0
365	Monensin (ppm)	0148	2.155	0.0100	2.115	0.2210	0.1917	16	0.18	365.03	0
365	Monensin (ppm)	0610	2.185	0.1300	2.115	0.2210	0.1917	16	0.32	365.02	0
365	Monensin (ppm)	2113	2.300	0.4000	2.115	0.2210	0.1917	16	0.84	365.05	0
365	Monensin (ppm)	0148	2.362	0.0610	2.115	0.2210	0.1917	16	1.11	365.04	0
365	Monensin (ppm)	0619	2.405	0.4300	2.115	0.2210	0.1917	16	1.31	365.05	0
365	Monensin (ppm)	0036	2.550	0.7000	2.115	0.2210	0.1917	16	1.97	365.03	0
370	Novobiocin (ppm)	0218	< 0.3					0		370.03	5
373	Oxytetracycline (ppm)	0939	1.282	0.0210				3		373.06	0
373	Oxytetracycline (ppm)	0619	2.025	0.5700				3		373.06	0
373	Oxytetracycline (ppm)	0148	2.223	0.3780				3		373.06	0
373	Oxytetracycline (ppm)	0160	7.235	0.7700				3		373.06	2
373	Oxytetracycline (ppm)	0218	< 2					3		373.03	5
379	Salinomycin (ppm)	0218	< 0.1					0		379.05	5
382	Sulfamethazine (ppm)	0218	1.955	0.0300				2		382.02	0
382	Sulfamethazine (ppm)	0148	2.066	0.0630				2		382.04	0
382	Sulfamethazine (ppm)	2053	< 1					2		382.03	5
388	Tylosin (ppm)	0148	0.2150	0.0640				3		388.05	0
388	Tylosin (ppm)	0160	0.3900	0.0600				3		388.05	0
388	Tylosin (ppm)	0619	0.4700	0.0800				3		388.05	0
388	Tylosin (ppm)	0218	< 0.1					3		388.05	5
388	Tylosin (ppm)	0875	< 1					3		388.05	5
389	Virginiamycin (ppm)	0218	< 0.1					0		389.03	5
391	Narasin (ppm)	0218	< 0.1					0		391.03	5
400	Water Activity (Units)	0843	0.5145	0.0030	0.5378	0.0175	0.0056	10	-1.33	400.99	0
400	Water Activity (Units)	0843	0.5171	0.0018	0.5378	0.0175	0.0056	10	-1.18	400.01	0
400	Water Activity (Units)	2168	0.5260	0.0180	0.5378	0.0175	0.0056	10	-0.67	400.01	0
400	Water Activity (Units)	0942	0.5340	0.0020	0.5378	0.0175	0.0056	10	-0.22	400.01	0
400	Water Activity (Units)	0407	0.5396	0.0024	0.5378	0.0175	0.0056	10	0.10	400.01	0
400	Water Activity (Units)	2174	0.5405	0.0130	0.5378	0.0175	0.0056	10	0.16	400.01	0
400	Water Activity (Units)	0589	0.5430	0.0060	0.5378	0.0175	0.0056	10	0.30	400.01	0
400	Water Activity (Units)	2033	0.5445	0.0070	0.5378	0.0175	0.0056	10	0.38	400.99	0
400	Water Activity (Units)	2181	0.5576	0.0008	0.5378	0.0175	0.0056	10	1.13	400.01	0
400	Water Activity (Units)	2073	0.5610	0.0020	0.5378	0.0175	0.0056	10	1.33	400.99	0
400	Water Activity (Units)	2109	0.3327	0.5440	0.5378	0.0175	0.0056	10	-11.73	400.01	1
516	Arsenic, Total (ppm)	0553	0.2655	0.0150	0.3674	0.0821	0.0182	9	-1.24	516.53	0
516	Arsenic, Total (ppm)	0199	0.2750	0.0100	0.3674	0.0821	0.0182	9	-1.13	516.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
516	Arsenic, Total (ppm)	0171	0.3235	0.0110	0.3674	0.0821	0.0182	9	-0.53	516.00	0
516	Arsenic, Total (ppm)	0939	0.3400	0.0200	0.3674	0.0821	0.0182	9	-0.33	516.53	0
516	Arsenic, Total (ppm)	2053	0.3650	0.0100	0.3674	0.0821	0.0182	9	-0.03	516.53	0
516	Arsenic, Total (ppm)	0910	0.4100	0.0200	0.3674	0.0821	0.0182	9	0.52	516.52	0
516	Arsenic, Total (ppm)	0227	0.4105	0.0350	0.3674	0.0821	0.0182	9	0.52	516.53	0
516	Arsenic, Total (ppm)	0918	0.4552	0.0132	0.3674	0.0821	0.0182	9	1.07	516.53	0
516	Arsenic, Total (ppm)	2012	0.4620	0.0300	0.3674	0.0821	0.0182	9	1.15	516.53	0
516	Arsenic, Total (ppm)	0407	1.145	0.3330	0.3674	0.0821	0.0182	9	9.46	516.43	1
516	Arsenic, Total (ppm)	0682	< 10		0.3674	0.0821	0.0182	9		516.43	5
518	Cadmium (ppm)	0171	0.0875	0.0010	0.1106	0.0191	0.0047	11	-1.21	518.41	0
518	Cadmium (ppm)	0939	0.0910	0.0020	0.1106	0.0191	0.0047	11	-1.02	518.53	0
518	Cadmium (ppm)	0199	0.0920	0.0020	0.1106	0.0191	0.0047	11	-0.97	518.53	0
518	Cadmium (ppm)	0407	0.1054	0.0073	0.1106	0.0191	0.0047	11	-0.27	518.43	0
518	Cadmium (ppm)	0553	0.1055	0.0070	0.1106	0.0191	0.0047	11	-0.27	518.53	0
518	Cadmium (ppm)	0918	0.1082	0.0054	0.1106	0.0191	0.0047	11	-0.12	518.53	0
518	Cadmium (ppm)	2012	0.1125	0.0130	0.1106	0.0191	0.0047	11	0.10	518.53	0
518	Cadmium (ppm)	0227	0.1180	0.0120	0.1106	0.0191	0.0047	11	0.39	518.53	0
518	Cadmium (ppm)	0968	0.1270	0.0020	0.1106	0.0191	0.0047	11	0.86	518.43	0
518	Cadmium (ppm)	0910	0.1300	0.0000	0.1106	0.0191	0.0047	11	1.01	518.52	0
518	Cadmium (ppm)	0682	0.3310	0.0000	0.1106	0.0191	0.0047	11	11.51	518.43	0
518	Cadmium (ppm)	0407	0.1275	0.0311	0.1106	0.0191	0.0047	11	0.88	518.41	1
520	Chromium (ppm)	0407	1.132	0.0208	10.14	6.757	0.3570	11	-1.33	520.43	0
520	Chromium (ppm)	0918	4.153	0.8798	10.14	6.757	0.3570	11	-0.89	520.53	0
520	Chromium (ppm)	0171	4.483	0.1010	10.14	6.757	0.3570	11	-0.84	520.41	0
520	Chromium (ppm)	0968	4.805	0.0370	10.14	6.757	0.3570	11	-0.79	520.43	0
520	Chromium (ppm)	0407	6.200	1.117	10.14	6.757	0.3570	11	-0.58	520.41	0
520	Chromium (ppm)	0939	13.33	0.2800	10.14	6.757	0.3570	11	0.47	520.53	0
520	Chromium (ppm)	0693	13.47	0.6718	10.14	6.757	0.3570	11	0.49	520.42	0
520	Chromium (ppm)	0164	14.90	0.2000	10.14	6.757	0.3570	11	0.70	520.53	0
520	Chromium (ppm)	0553	15.20	0.6000	10.14	6.757	0.3570	11	0.75	520.53	0
520	Chromium (ppm)	0682	16.45	0.0000	10.14	6.757	0.3570	11	0.93	520.43	0
520	Chromium (ppm)	0510	17.40	0.0200	10.14	6.757	0.3570	11	1.07	520.43	0
526	Lead (ppm)	0171	0.1595	0.0030	0.2819	0.0363	0.0308	12	-3.37	526.41	0
526	Lead (ppm)	0199	0.2300	0.0000	0.2819	0.0363	0.0308	12	-1.43	526.53	0
526	Lead (ppm)	0939	0.2550	0.0300	0.2819	0.0363	0.0308	12	-0.74	526.53	0
526	Lead (ppm)	0553	0.2655	0.0270	0.2819	0.0363	0.0308	12	-0.45	526.53	0
526	Lead (ppm)	0227	0.2820	0.0160	0.2819	0.0363	0.0308	12	0.00	526.53	0
526	Lead (ppm)	2053	0.2850	0.0300	0.2819	0.0363	0.0308	12	0.08	526.53	0
526	Lead (ppm)	0968	0.2875	0.0010	0.2819	0.0363	0.0308	12	0.15	526.43	0
526	Lead (ppm)	0918	0.2954	0.0082	0.2819	0.0363	0.0308	12	0.37	526.53	0
526	Lead (ppm)	0407	0.3005	0.0897	0.2819	0.0363	0.0308	12	0.51	526.43	0
526	Lead (ppm)	0910	0.3100	0.0200	0.2819	0.0363	0.0308	12	0.77	526.52	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
526	Lead (ppm)	0407	0.3212	0.1216	0.2819	0.0363	0.0308	12	1.08	526.41	0
526	Lead (ppm)	2012	0.3235	0.0230	0.2819	0.0363	0.0308	12	1.14	526.53	0
526	Lead (ppm)	0682	< 3		0.2819	0.0363	0.0308	12		526.43	5
529	Mercury (ppb)	2012	2.658	0.1510				3		529.99	0
529	Mercury (ppb)	0199	4.000	0.0000				3		529.99	0
529	Mercury (ppb)	0939	17.00	6.000				3		529.99	0
529	Mercury (ppb)	0227	< 10					3		529.99	5
529	Mercury (ppb)	0682	< 10000					3		529.99	5
539	Nickel (ppm)	0918	2.616	0.1713	5.848	1.712	0.2231	7	-1.89	539.53	0
539	Nickel (ppm)	0171	4.378	0.0640	5.848	1.712	0.2231	7	-0.86	539.41	0
539	Nickel (ppm)	0407	5.952	0.6290	5.848	1.712	0.2231	7	0.06	539.41	0
539	Nickel (ppm)	0407	6.029	0.2676	5.848	1.712	0.2231	7	0.11	539.43	0
539	Nickel (ppm)	0939	6.765	0.2700	5.848	1.712	0.2231	7	0.54	539.53	0
539	Nickel (ppm)	0682	7.180	0.0000	5.848	1.712	0.2231	7	0.78	539.43	0
539	Nickel (ppm)	0553	7.350	0.1600	5.848	1.712	0.2231	7	0.88	539.53	0
702	Butyric Acid (4:0) (%)	0619	0.0000	0.0000				0		702.00	4
704	Caproic Acid (6:0) (%)	0619	0.0000	0.0000				0		704.00	4
706	Caprylic acid (8:0) (%) (w/w)	0619	0.0000	0.0000				0		706.01	4
708	Capric acid (10:0) (%) (w/w)	0619	0.0000	0.0000				0		708.01	4
710	Lauric Acid (12:0) (%) (w/w)	0676	0.0015	0.0010				1		710.99	0
710	Lauric Acid (12:0) (%) (w/w)	0619	0.0000	0.0000				1		710.01	4
710	Lauric Acid (12:0) (%) (w/w)	0226	< 0.005					1		710.99	5
710	Lauric Acid (12:0) (%) (w/w)	0164	< 0.02					1		710.99	5
714	Myristic Acid (14:0) (%) (w/w)	0226	0.0130	0.0020				2		714.99	0
714	Myristic Acid (14:0) (%) (w/w)	0619	0.0240	0.0020				2		714.01	0
714	Myristic Acid (14:0) (%) (w/w)	0164	< 0.02					2		714.99	5
716	Palmitic Acid (16:0) (%) (w/w)	0226	0.6975	0.0030				3		716.99	0
716	Palmitic Acid (16:0) (%) (w/w)	0164	0.7850	0.0100				3		716.99	0
716	Palmitic Acid (16:0) (%) (w/w)	0619	0.7940	0.0120				3		716.01	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0676	0.0200	0.0000				3		718.99	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0226	0.0235	0.0030				3		718.99	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0619	0.0315	0.0030				3		718.01	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0164	< 0.04					3		718.99	5
722	Stearic Acid (18:0) (%) (w/w)	0619	0.1215	0.0010				3		722.01	0
722	Stearic Acid (18:0) (%) (w/w)	0226	0.1535	0.0030				3		722.99	0
722	Stearic Acid (18:0) (%) (w/w)	0164	0.1600	0.0000				3		722.99	0
724	Oleic Acid (9c-18:1) (%) (w/w)	0226	1.110	0.0230				3		724.99	0
724	Oleic Acid (9c-18:1) (%) (w/w)	0164	1.120	0.0000				3		724.99	0
724	Oleic Acid (9c-18:1) (%) (w/w)	0619	1.205	0.0300				3		724.01	0
726	Linoleic Acid (9c,12c-18:2) (%) (w/w)	0226	1.830	0.0000				3		726.99	0
726	Linoleic Acid (9c,12c-18:2) (%) (w/w)	0164	1.920	0.0200				3		726.99	0
726	Linoleic Acid (9c,12c-18:2) (%) (w/w)	0676	1.940	0.0030				3		726.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0619	2.725	0.0100				3	726.01	2	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/w))	0226	0.1210	0.0000				3	728.99	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/w))	0676	0.1260	0.0000				3	728.99	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/w))	0164	0.1300	0.0000				3	728.99	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/w))	0619	0.0985	0.0030				3	728.01	1	
730	Arachidic Acid (20:0) (% (w/w))	0226	0.0125	0.0010				2	730.99	0	
730	Arachidic Acid (20:0) (% (w/w))	0619	0.0270	0.0020				2	730.01	0	
730	Arachidic Acid (20:0) (% (w/w))	0164	< 0.02					2	730.99	5	
732	Gondoic Acid (11c-20:1) (% (w/w))	0226	0.0200	0.0000				2	732.99	0	
732	Gondoic Acid (11c-20:1) (% (w/w))	0164	0.0300	0.0000				2	732.99	0	
732	Gondoic Acid (11c-20:1) (% (w/w))	0619	0.0000	0.0000				2	732.01	4	
736	Arachidonic Acid (5c,8c,11c,14c-20:4) (% (w/w))	0676	0.0020	0.0000				1	736.99	0	
736	Arachidonic Acid (5c,8c,11c,14c-20:4) (% (w/w))	0619	0.0000	0.0000				1	736.01	4	
738	Mead Acid (11c,14c,17c-20:3) (% (w/w))	0619	0.0000	0.0000				0	738.01	4	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c,20:5) (% (w/w))	0619	0.0000	0.0000				0	740.01	4	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c,20:5) (% (w/w))	0226	0.0000	0.0000				0	740.99	4	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c,20:5) (% (w/w))	0676	0.0000	0.0000				0	740.99	4	
742	Behenic Acid (22:0) (% (w/w))	0226	0.0085	0.0010				1	742.99	0	
742	Behenic Acid (22:0) (% (w/w))	0164	< 0.02					1	742.99	5	
744	Erucic Acid (13c-22:1) (% (w/w))	0619	0.0000	0.0000				0	744.01	4	
744	Erucic Acid (13c-22:1) (% (w/w))	0226	< 0.005					0	744.99	5	
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c,22:5) (% (w/w))	0619	0.0000	0.0000				0	746.01	4	
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c,22:5) (% (w/w))	0226	0.0000	0.0000				0	746.99	4	
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c,22:5) (% (w/w))	0676	0.0000	0.0000				0	746.99	4	
748	Lignoceric Acid (24:0) (% (w/w))	0226	0.0115	0.0010				1	748.99	0	
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22:6) (% (w/w))	0619	0.0000	0.0000				0	750.01	4	
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22:6) (% (w/w))	0226	0.0000	0.0000				0	750.99	4	
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22:6) (% (w/w))	0676	0.0000	0.0000				0	750.99	4	
752	Nervonic Acid (24:1) isomers (% (w/w))	0619	0.0000	0.0000				0	752.01	4	
752	Nervonic Acid (24:1) isomers (% (w/w))	0226	0.0000	0.0000				0	752.99	4	
754	Total n-3 Polyunsaturated (Omega-3) Fatty Acids (% (w/w))	0676	0.1300	0.0000				2	754.99	0	
754	Total n-3 Polyunsaturated (Omega-3) Fatty Acids (% (w/w))	0164	0.1350	0.0100				2	754.99	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Acids (% (w/w))	0164	1.925	0.0300				2	756.99	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Acids (% (w/w))	0676	1.950	0.0000				2	756.99	0	
758	Total Saturated Fatty Acids (% (w/w))	0164	1.025	0.0100				1	758.99	0	
762	Total Monounsaturated Fatty Acids (% (w/w))	0164	1.180	0.0000				1	762.99	0	
766	Total Polyunsaturated Fatty Acids (% (w/w))	0164	2.065	0.0300				1	766.99	0	
770	Total Fat (equivalent to NLEA) (% (w/w))	0164	4.490	0.0200				1	770.99	0	
772	Total Fatty Acids (% (w/w))	0226	4.000	0.0200				2	772.99	0	
772	Total Fatty Acids (% (w/w))	0164	4.295	0.0100				2	772.99	0	

Notes: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2 and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = Rejected for duplicates too far apart, 2 = Rejected as outlier, 5 = A reporting limit and 4 = zeros submitted as

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			

values. Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). Flag 3 indicates not used in statistics.