

Animal Feed Scheme

Dry Cat Feed

Test Material Code # 202023

Labs Reporting: 190

Analytes Reported 111

Issue Date : 04/30/2020

Analyte Proficiency Testing Report

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
000	Urea (%)	0278	0.2000	0.0000				1		000.02	0
001	Loss on Drying (%)	0618	5.425	0.0300	7.473	0.2384	0.1028	79	-8.59	001.99	0
001	Loss on Drying (%)	2155	6.445	0.4100	7.473	0.2384	0.1028	79	-4.31	001.99	0
001	Loss on Drying (%)	0019	6.700	0.4000	7.473	0.2384	0.1028	79	-3.24	001.07	0
001	Loss on Drying (%)	0536	6.870	0.0200	7.473	0.2384	0.1028	79	-2.53	001.99	0
001	Loss on Drying (%)	0038	6.935	0.3300	7.473	0.2384	0.1028	79	-2.26	001.07	0
001	Loss on Drying (%)	0034	6.990	0.0800	7.473	0.2384	0.1028	79	-2.03	001.07	0
001	Loss on Drying (%)	0297	6.990	0.2600	7.473	0.2384	0.1028	79	-2.03	001.07	0
001	Loss on Drying (%)	0845	7.010	0.1600	7.473	0.2384	0.1028	79	-1.94	001.07	0
001	Loss on Drying (%)	0889	7.050	0.0000	7.473	0.2384	0.1028	79	-1.77	001.99	0
001	Loss on Drying (%)	2081	7.060	0.0600	7.473	0.2384	0.1028	79	-1.73	001.99	0
001	Loss on Drying (%)	2252	7.080	0.0000	7.473	0.2384	0.1028	79	-1.65	001.07	0
001	Loss on Drying (%)	2259	7.097	0.0840	7.473	0.2384	0.1028	79	-1.58	001.07	0
001	Loss on Drying (%)	0676	7.115	0.0300	7.473	0.2384	0.1028	79	-1.50	001.99	0
001	Loss on Drying (%)	0510	7.150	0.1000	7.473	0.2384	0.1028	79	-1.36	001.99	0
001	Loss on Drying (%)	2124	7.245	0.0700	7.473	0.2384	0.1028	79	-0.96	001.03	0
001	Loss on Drying (%)	2220	7.295	0.1500	7.473	0.2384	0.1028	79	-0.75	001.07	0
001	Loss on Drying (%)	0345	7.300	0.0400	7.473	0.2384	0.1028	79	-0.73	001.07	0
001	Loss on Drying (%)	2187	7.310	0.0200	7.473	0.2384	0.1028	79	-0.68	001.07	0
001	Loss on Drying (%)	0918	7.350	0.0600	7.473	0.2384	0.1028	79	-0.52	001.99	0
001	Loss on Drying (%)	2136	7.365	0.1500	7.473	0.2384	0.1028	79	-0.45	001.99	0
001	Loss on Drying (%)	2009	7.372	0.0006	7.473	0.2384	0.1028	79	-0.43	001.07	0
001	Loss on Drying (%)	2115	7.380	0.0200	7.473	0.2384	0.1028	79	-0.39	001.99	0
001	Loss on Drying (%)	0089	7.390	0.0000	7.473	0.2384	0.1028	79	-0.35	001.07	0
001	Loss on Drying (%)	0169	7.395	0.0100	7.473	0.2384	0.1028	79	-0.33	001.00	0
001	Loss on Drying (%)	0948	7.395	0.0100	7.473	0.2384	0.1028	79	-0.33	001.99	0
001	Loss on Drying (%)	0853	7.400	0.2000	7.473	0.2384	0.1028	79	-0.31	001.07	0
001	Loss on Drying (%)	2012	7.430	0.0000	7.473	0.2384	0.1028	79	-0.18	001.07	0
001	Loss on Drying (%)	0066	7.445	0.1100	7.473	0.2384	0.1028	79	-0.12	001.07	0
001	Loss on Drying (%)	0843	7.455	0.2900	7.473	0.2384	0.1028	79	-0.08	001.03	0
001	Loss on Drying (%)	0683	7.455	0.1100	7.473	0.2384	0.1028	79	-0.08	001.07	0
001	Loss on Drying (%)	0843	7.455	0.2900	7.473	0.2384	0.1028	79	-0.08	001.07	0
001	Loss on Drying (%)	0644	7.460	0.0600	7.473	0.2384	0.1028	79	-0.06	001.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
001	Loss on Drying (%)	0139	7.460	0.0000	7.473	0.2384	0.1028	79	-0.06	001.07	0
001	Loss on Drying (%)	0142	7.460	0.1000	7.473	0.2384	0.1028	79	-0.06	001.07	0
001	Loss on Drying (%)	2156	7.480	0.2800	7.473	0.2384	0.1028	79	0.03	001.07	0
001	Loss on Drying (%)	0083	7.485	0.0100	7.473	0.2384	0.1028	79	0.05	001.07	0
001	Loss on Drying (%)	2059	7.485	0.0100	7.473	0.2384	0.1028	79	0.05	001.07	0
001	Loss on Drying (%)	2219	7.495	0.1900	7.473	0.2384	0.1028	79	0.09	001.07	0
001	Loss on Drying (%)	2242	7.495	0.0900	7.473	0.2384	0.1028	79	0.09	001.99	0
001	Loss on Drying (%)	0876	7.500	0.0000	7.473	0.2384	0.1028	79	0.11	001.07	0
001	Loss on Drying (%)	0910	7.500	0.2000	7.473	0.2384	0.1028	79	0.11	001.99	0
001	Loss on Drying (%)	2264	7.504	0.0060	7.473	0.2384	0.1028	79	0.13	001.07	0
001	Loss on Drying (%)	0226	7.505	0.0700	7.473	0.2384	0.1028	79	0.13	001.07	0
001	Loss on Drying (%)	0868	7.510	0.0600	7.473	0.2384	0.1028	79	0.15	001.03	0
001	Loss on Drying (%)	0015	7.510	0.1200	7.473	0.2384	0.1028	79	0.15	001.07	0
001	Loss on Drying (%)	2192	7.550	0.0000	7.473	0.2384	0.1028	79	0.32	001.07	0
001	Loss on Drying (%)	0571	7.550	0.1200	7.473	0.2384	0.1028	79	0.32	001.07	0
001	Loss on Drying (%)	0619	7.555	0.0100	7.473	0.2384	0.1028	79	0.34	001.03	0
001	Loss on Drying (%)	2268	7.560	0.3800	7.473	0.2384	0.1028	79	0.36	001.99	0
001	Loss on Drying (%)	0610	7.575	0.0100	7.473	0.2384	0.1028	79	0.43	001.99	0
001	Loss on Drying (%)	2053	7.580	0.0200	7.473	0.2384	0.1028	79	0.45	001.07	0
001	Loss on Drying (%)	0353	7.585	0.0100	7.473	0.2384	0.1028	79	0.47	001.07	0
001	Loss on Drying (%)	2293	7.590	0.2600	7.473	0.2384	0.1028	79	0.49	001.07	0
001	Loss on Drying (%)	0309	7.591	0.1270	7.473	0.2384	0.1028	79	0.49	001.00	0
001	Loss on Drying (%)	2295	7.595	0.0100	7.473	0.2384	0.1028	79	0.51	001.03	0
001	Loss on Drying (%)	0590	7.595	0.1700	7.473	0.2384	0.1028	79	0.51	001.07	0
001	Loss on Drying (%)	0904	7.595	0.3700	7.473	0.2384	0.1028	79	0.51	001.99	0
001	Loss on Drying (%)	0689	7.600	0.0000	7.473	0.2384	0.1028	79	0.53	001.07	0
001	Loss on Drying (%)	0035	7.605	0.2100	7.473	0.2384	0.1028	79	0.55	001.07	0
001	Loss on Drying (%)	0037	7.620	0.0400	7.473	0.2384	0.1028	79	0.62	001.99	0
001	Loss on Drying (%)	0065	7.620	0.0000	7.473	0.2384	0.1028	79	0.62	001.07	0
001	Loss on Drying (%)	0878	7.630	0.0400	7.473	0.2384	0.1028	79	0.66	001.03	0
001	Loss on Drying (%)	0581	7.635	0.0500	7.473	0.2384	0.1028	79	0.68	001.07	0
001	Loss on Drying (%)	0033	7.640	0.0400	7.473	0.2384	0.1028	79	0.70	001.07	0
001	Loss on Drying (%)	0638	7.650	0.0200	7.473	0.2384	0.1028	79	0.74	001.07	0
001	Loss on Drying (%)	0003	7.660	0.0600	7.473	0.2384	0.1028	79	0.78	001.07	0
001	Loss on Drying (%)	0098	7.680	0.0400	7.473	0.2384	0.1028	79	0.87	001.07	0
001	Loss on Drying (%)	0186	7.680	0.0940	7.473	0.2384	0.1028	79	0.87	001.99	0
001	Loss on Drying (%)	0407	7.688	0.0362	7.473	0.2384	0.1028	79	0.90	001.07	0
001	Loss on Drying (%)	0693	7.703	0.1480	7.473	0.2384	0.1028	79	0.96	001.07	0
001	Loss on Drying (%)	0148	7.705	0.1300	7.473	0.2384	0.1028	79	0.97	001.07	0
001	Loss on Drying (%)	0872	7.745	0.0100	7.473	0.2384	0.1028	79	1.14	001.07	0
001	Loss on Drying (%)	0019	7.750	0.2200	7.473	0.2384	0.1028	79	1.16	001.99	0
001	Loss on Drying (%)	2109	7.785	0.3300	7.473	0.2384	0.1028	79	1.31	001.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
001	Loss on Drying (%)	0366	7.800	0.2000	7.473	0.2384	0.1028	79	1.37	001.07	0
001	Loss on Drying (%)	0357	7.851	0.0062	7.473	0.2384	0.1028	79	1.58	001.99	0
001	Loss on Drying (%)	0968	7.905	0.1800	7.473	0.2384	0.1028	79	1.81	001.05	0
001	Loss on Drying (%)	0278	7.945	0.0900	7.473	0.2384	0.1028	79	1.98	001.07	0
001	Loss on Drying (%)	0682	8.390	0.0000	7.473	0.2384	0.1028	79	3.85	001.07	0
001	Loss on Drying (%)	0045	8.085	1.290	7.473	0.2384	0.1028	79	2.57	001.07	1
002	Protein, Crude (%)	0062	30.13	0.0020	31.94	0.3827	0.1856	187	-4.73	002.06	0
002	Protein, Crude (%)	2073	30.26	0.4500	31.94	0.3827	0.1856	187	-4.40	002.05	0
002	Protein, Crude (%)	0297	30.28	0.1600	31.94	0.3827	0.1856	187	-4.33	002.11	0
002	Protein, Crude (%)	0553	30.33	0.1800	31.94	0.3827	0.1856	187	-4.20	002.11	0
002	Protein, Crude (%)	2268	30.45	0.3700	31.94	0.3827	0.1856	187	-3.90	002.06	0
002	Protein, Crude (%)	0536	30.55	0.0500	31.94	0.3827	0.1856	187	-3.64	002.05	0
002	Protein, Crude (%)	2006	31.14	0.1700	31.94	0.3827	0.1856	187	-2.10	002.05	0
002	Protein, Crude (%)	2314	31.15	0.1000	31.94	0.3827	0.1856	187	-2.06	002.01	0
002	Protein, Crude (%)	2193	31.18	0.1000	31.94	0.3827	0.1856	187	-1.98	002.01	0
002	Protein, Crude (%)	0939	31.22	0.0900	31.94	0.3827	0.1856	187	-1.89	002.05	0
002	Protein, Crude (%)	0626	31.24	0.2200	31.94	0.3827	0.1856	187	-1.82	002.06	0
002	Protein, Crude (%)	2188	31.31	0.0100	31.94	0.3827	0.1856	187	-1.65	002.01	0
002	Protein, Crude (%)	2022	31.33	0.0200	31.94	0.3827	0.1856	187	-1.59	002.05	0
002	Protein, Crude (%)	2259	31.34	0.0320	31.94	0.3827	0.1856	187	-1.56	002.01	0
002	Protein, Crude (%)	0619	31.40	0.2000	31.94	0.3827	0.1856	187	-1.41	002.01	0
002	Protein, Crude (%)	0619	31.40	0.2000	31.94	0.3827	0.1856	187	-1.41	002.05	0
002	Protein, Crude (%)	0674	31.40	0.0000	31.94	0.3827	0.1856	187	-1.41	002.05	0
002	Protein, Crude (%)	0689	31.40	0.2000	31.94	0.3827	0.1856	187	-1.41	002.05	0
002	Protein, Crude (%)	2022	31.45	0.0310	31.94	0.3827	0.1856	187	-1.28	002.06	0
002	Protein, Crude (%)	0520	31.45	0.1000	31.94	0.3827	0.1856	187	-1.27	002.06	0
002	Protein, Crude (%)	2009	31.47	0.0067	31.94	0.3827	0.1856	187	-1.23	002.05	0
002	Protein, Crude (%)	2295	31.49	0.0100	31.94	0.3827	0.1856	187	-1.18	002.05	0
002	Protein, Crude (%)	2304	31.52	0.3710	31.94	0.3827	0.1856	187	-1.10	002.01	0
002	Protein, Crude (%)	0653	31.52	0.0000	31.94	0.3827	0.1856	187	-1.09	002.01	0
002	Protein, Crude (%)	0598	31.52	0.1800	31.94	0.3827	0.1856	187	-1.09	002.06	0
002	Protein, Crude (%)	0065	31.52	0.0395	31.94	0.3827	0.1856	187	-1.09	002.06	0
002	Protein, Crude (%)	2109	31.54	0.1900	31.94	0.3827	0.1856	187	-1.05	002.05	0
002	Protein, Crude (%)	0910	31.55	0.3000	31.94	0.3827	0.1856	187	-1.01	002.01	0
002	Protein, Crude (%)	0265	31.55	0.3000	31.94	0.3827	0.1856	187	-1.01	002.06	0
002	Protein, Crude (%)	0294	31.55	0.1000	31.94	0.3827	0.1856	187	-1.01	002.06	0
002	Protein, Crude (%)	0144	31.56	0.0700	31.94	0.3827	0.1856	187	-1.00	002.06	0
002	Protein, Crude (%)	0870	31.56	0.3660	31.94	0.3827	0.1856	187	-0.99	002.01	0
002	Protein, Crude (%)	0638	31.56	0.6200	31.94	0.3827	0.1856	187	-0.99	002.04	0
002	Protein, Crude (%)	2293	31.57	0.1600	31.94	0.3827	0.1856	187	-0.96	002.06	0
002	Protein, Crude (%)	2156	31.57	0.2570	31.94	0.3827	0.1856	187	-0.96	002.06	0
002	Protein, Crude (%)	0860	31.58	0.0000	31.94	0.3827	0.1856	187	-0.95	002.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0345	31.58	0.5500	31.94	0.3827	0.1856	187	-0.95	002.06	0
002	Protein, Crude (%)	0675	31.58	0.0400	31.94	0.3827	0.1856	187	-0.94	002.05	0
002	Protein, Crude (%)	0939	31.59	0.0900	31.94	0.3827	0.1856	187	-0.92	002.06	0
002	Protein, Crude (%)	2196	31.59	0.0000	31.94	0.3827	0.1856	187	-0.91	002.01	0
002	Protein, Crude (%)	0644	31.60	0.3820	31.94	0.3827	0.1856	187	-0.89	002.06	0
002	Protein, Crude (%)	0354	31.61	0.1000	31.94	0.3827	0.1856	187	-0.86	002.05	0
002	Protein, Crude (%)	0018	31.61	0.1000	31.94	0.3827	0.1856	187	-0.86	002.06	0
002	Protein, Crude (%)	2122	31.63	0.0000	31.94	0.3827	0.1856	187	-0.80	002.04	0
002	Protein, Crude (%)	0941	31.63	0.2000	31.94	0.3827	0.1856	187	-0.80	002.05	0
002	Protein, Crude (%)	2053	31.65	0.1000	31.94	0.3827	0.1856	187	-0.75	002.06	0
002	Protein, Crude (%)	2220	31.66	0.0200	31.94	0.3827	0.1856	187	-0.73	002.06	0
002	Protein, Crude (%)	2219	31.70	0.3700	31.94	0.3827	0.1856	187	-0.63	002.06	0
002	Protein, Crude (%)	2136	31.70	0.2000	31.94	0.3827	0.1856	187	-0.62	002.08	0
002	Protein, Crude (%)	0066	31.73	0.1000	31.94	0.3827	0.1856	187	-0.54	002.02	0
002	Protein, Crude (%)	0683	31.74	0.1500	31.94	0.3827	0.1856	187	-0.53	002.05	0
002	Protein, Crude (%)	0175	31.75	0.1000	31.94	0.3827	0.1856	187	-0.49	002.06	0
002	Protein, Crude (%)	0692	31.75	0.1000	31.94	0.3827	0.1856	187	-0.49	002.06	0
002	Protein, Crude (%)	0876	31.75	0.5000	31.94	0.3827	0.1856	187	-0.49	002.06	0
002	Protein, Crude (%)	0884	31.75	0.3000	31.94	0.3827	0.1856	187	-0.49	002.06	0
002	Protein, Crude (%)	0098	31.76	0.0300	31.94	0.3827	0.1856	187	-0.48	002.08	0
002	Protein, Crude (%)	0872	31.77	0.3510	31.94	0.3827	0.1856	187	-0.44	002.06	0
002	Protein, Crude (%)	0581	31.77	0.4600	31.94	0.3827	0.1856	187	-0.44	002.06	0
002	Protein, Crude (%)	2264	31.77	0.1080	31.94	0.3827	0.1856	187	-0.43	002.05	0
002	Protein, Crude (%)	0859	31.78	0.1150	31.94	0.3827	0.1856	187	-0.42	002.06	0
002	Protein, Crude (%)	2187	31.78	0.1200	31.94	0.3827	0.1856	187	-0.41	002.06	0
002	Protein, Crude (%)	0968	31.79	0.4600	31.94	0.3827	0.1856	187	-0.38	002.01	0
002	Protein, Crude (%)	2115	31.80	0.0400	31.94	0.3827	0.1856	187	-0.36	002.05	0
002	Protein, Crude (%)	2303	31.81	0.3600	31.94	0.3827	0.1856	187	-0.33	002.06	0
002	Protein, Crude (%)	0164	31.81	0.5000	31.94	0.3827	0.1856	187	-0.33	002.01	0
002	Protein, Crude (%)	0868	31.81	0.1600	31.94	0.3827	0.1856	187	-0.33	002.05	0
002	Protein, Crude (%)	0897	31.82	0.0000	31.94	0.3827	0.1856	187	-0.31	002.05	0
002	Protein, Crude (%)	0035	31.82	0.0600	31.94	0.3827	0.1856	187	-0.31	002.06	0
002	Protein, Crude (%)	0036	31.82	0.0255	31.94	0.3827	0.1856	187	-0.30	002.02	0
002	Protein, Crude (%)	0100	31.83	0.0800	31.94	0.3827	0.1856	187	-0.28	002.06	0
002	Protein, Crude (%)	2207	31.84	0.1300	31.94	0.3827	0.1856	187	-0.27	002.06	0
002	Protein, Crude (%)	0958	31.85	0.0300	31.94	0.3827	0.1856	187	-0.24	002.06	0
002	Protein, Crude (%)	2169	31.85	0.0900	31.94	0.3827	0.1856	187	-0.24	002.06	0
002	Protein, Crude (%)	0652	31.85	0.3000	31.94	0.3827	0.1856	187	-0.23	002.01	0
002	Protein, Crude (%)	0853	31.85	0.3000	31.94	0.3827	0.1856	187	-0.23	002.06	0
002	Protein, Crude (%)	0857	31.86	0.1500	31.94	0.3827	0.1856	187	-0.22	002.06	0
002	Protein, Crude (%)	2192	31.87	0.3100	31.94	0.3827	0.1856	187	-0.19	002.05	0
002	Protein, Crude (%)	0226	31.89	0.3700	31.94	0.3827	0.1856	187	-0.14	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	2252	31.89	0.0630	31.94	0.3827	0.1856	187	-0.13	002.06	0
002	Protein, Crude (%)	2109	31.90	0.1900	31.94	0.3827	0.1856	187	-0.11	002.06	0
002	Protein, Crude (%)	0610	31.90	0.0000	31.94	0.3827	0.1856	187	-0.10	002.01	0
002	Protein, Crude (%)	2012	31.90	0.0800	31.94	0.3827	0.1856	187	-0.10	002.05	0
002	Protein, Crude (%)	0278	31.90	0.2000	31.94	0.3827	0.1856	187	-0.10	002.06	0
002	Protein, Crude (%)	0510	31.90	0.2000	31.94	0.3827	0.1856	187	-0.10	002.06	0
002	Protein, Crude (%)	0956	31.90	0.2000	31.94	0.3827	0.1856	187	-0.10	002.06	0
002	Protein, Crude (%)	0541	31.92	0.1100	31.94	0.3827	0.1856	187	-0.06	002.06	0
002	Protein, Crude (%)	0529	31.93	0.1300	31.94	0.3827	0.1856	187	-0.03	002.06	0
002	Protein, Crude (%)	0646	31.94	0.3500	31.94	0.3827	0.1856	187	-0.01	002.06	0
002	Protein, Crude (%)	0878	31.95	0.0100	31.94	0.3827	0.1856	187	0.02	002.05	0
002	Protein, Crude (%)	2168	31.95	0.0900	31.94	0.3827	0.1856	187	0.02	002.06	0
002	Protein, Crude (%)	2152	31.95	0.0400	31.94	0.3827	0.1856	187	0.03	002.05	0
002	Protein, Crude (%)	0964	31.97	0.0260	31.94	0.3827	0.1856	187	0.07	002.06	0
002	Protein, Crude (%)	0051	31.97	0.5600	31.94	0.3827	0.1856	187	0.08	002.06	0
002	Protein, Crude (%)	0676	31.97	0.1800	31.94	0.3827	0.1856	187	0.08	002.06	0
002	Protein, Crude (%)	0914	32.00	0.0500	31.94	0.3827	0.1856	187	0.15	002.01	0
002	Protein, Crude (%)	2312	32.00	0.1500	31.94	0.3827	0.1856	187	0.15	002.04	0
002	Protein, Crude (%)	0574	32.00	0.2100	31.94	0.3827	0.1856	187	0.15	002.06	0
002	Protein, Crude (%)	0870	32.00	0.1019	31.94	0.3827	0.1856	187	0.15	002.06	0
002	Protein, Crude (%)	2291	32.00	0.3000	31.94	0.3827	0.1856	187	0.16	002.05	0
002	Protein, Crude (%)	0010	32.00	0.0000	31.94	0.3827	0.1856	187	0.16	002.06	0
002	Protein, Crude (%)	0123	32.00	0.0200	31.94	0.3827	0.1856	187	0.16	002.06	0
002	Protein, Crude (%)	0610	32.00	0.2000	31.94	0.3827	0.1856	187	0.16	002.06	0
002	Protein, Crude (%)	0619	32.00	0.2000	31.94	0.3827	0.1856	187	0.16	002.06	0
002	Protein, Crude (%)	2232	32.00	0.2000	31.94	0.3827	0.1856	187	0.16	002.06	0
002	Protein, Crude (%)	0693	32.00	0.1500	31.94	0.3827	0.1856	187	0.17	002.06	0
002	Protein, Crude (%)	0263	32.01	0.1400	31.94	0.3827	0.1856	187	0.19	002.06	0
002	Protein, Crude (%)	2124	32.01	0.0800	31.94	0.3827	0.1856	187	0.19	002.06	0
002	Protein, Crude (%)	2016	32.02	0.3900	31.94	0.3827	0.1856	187	0.20	002.06	0
002	Protein, Crude (%)	2174	32.02	0.0700	31.94	0.3827	0.1856	187	0.20	002.06	0
002	Protein, Crude (%)	0019	32.02	0.0400	31.94	0.3827	0.1856	187	0.21	002.06	0
002	Protein, Crude (%)	0918	32.02	0.1600	31.94	0.3827	0.1856	187	0.21	002.06	0
002	Protein, Crude (%)	2181	32.03	0.5100	31.94	0.3827	0.1856	187	0.23	002.05	0
002	Protein, Crude (%)	0169	32.03	0.2100	31.94	0.3827	0.1856	187	0.23	002.00	0
002	Protein, Crude (%)	0298	32.03	0.0200	31.94	0.3827	0.1856	187	0.24	002.06	0
002	Protein, Crude (%)	2304	32.03	0.2600	31.94	0.3827	0.1856	187	0.24	002.06	0
002	Protein, Crude (%)	0038	32.04	0.1870	31.94	0.3827	0.1856	187	0.25	002.06	0
002	Protein, Crude (%)	0142	32.04	0.0000	31.94	0.3827	0.1856	187	0.27	002.06	0
002	Protein, Crude (%)	0948	32.05	0.0300	31.94	0.3827	0.1856	187	0.28	002.05	0
002	Protein, Crude (%)	0009	32.05	0.1000	31.94	0.3827	0.1856	187	0.29	002.06	0
002	Protein, Crude (%)	0309	32.05	0.5000	31.94	0.3827	0.1856	187	0.29	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0366	32.05	0.7000	31.94	0.3827	0.1856	187	0.29	002.06	0
002	Protein, Crude (%)	0202	32.05	0.5800	31.94	0.3827	0.1856	187	0.29	002.06	0
002	Protein, Crude (%)	0554	32.06	0.1500	31.94	0.3827	0.1856	187	0.31	002.06	0
002	Protein, Crude (%)	0675	32.06	0.0100	31.94	0.3827	0.1856	187	0.31	002.06	0
002	Protein, Crude (%)	0353	32.07	0.0500	31.94	0.3827	0.1856	187	0.33	002.06	0
002	Protein, Crude (%)	0354	32.07	0.0000	31.94	0.3827	0.1856	187	0.34	002.06	0
002	Protein, Crude (%)	2162	32.07	0.2200	31.94	0.3827	0.1856	187	0.34	002.06	0
002	Protein, Crude (%)	2129	32.08	0.0320	31.94	0.3827	0.1856	187	0.36	002.06	0
002	Protein, Crude (%)	0297	32.09	0.3100	31.94	0.3827	0.1856	187	0.38	002.06	0
002	Protein, Crude (%)	0848	32.09	0.3300	31.94	0.3827	0.1856	187	0.38	002.06	0
002	Protein, Crude (%)	0953	32.09	0.0400	31.94	0.3827	0.1856	187	0.40	002.06	0
002	Protein, Crude (%)	0968	32.10	0.2470	31.94	0.3827	0.1856	187	0.42	002.06	0
002	Protein, Crude (%)	0017	32.10	0.2000	31.94	0.3827	0.1856	187	0.42	002.06	0
002	Protein, Crude (%)	0098	32.10	0.2000	31.94	0.3827	0.1856	187	0.42	002.06	0
002	Protein, Crude (%)	0726	32.15	0.0300	31.94	0.3827	0.1856	187	0.54	002.06	0
002	Protein, Crude (%)	0845	32.15	0.1600	31.94	0.3827	0.1856	187	0.55	002.00	0
002	Protein, Crude (%)	0003	32.15	0.5000	31.94	0.3827	0.1856	187	0.55	002.06	0
002	Protein, Crude (%)	2033	32.15	0.2200	31.94	0.3827	0.1856	187	0.55	002.06	0
002	Protein, Crude (%)	0682	32.16	0.0000	31.94	0.3827	0.1856	187	0.58	002.06	0
002	Protein, Crude (%)	0034	32.17	0.0370	31.94	0.3827	0.1856	187	0.60	002.06	0
002	Protein, Crude (%)	0139	32.18	0.3900	31.94	0.3827	0.1856	187	0.62	002.06	0
002	Protein, Crude (%)	0015	32.18	0.1200	31.94	0.3827	0.1856	187	0.63	002.06	0
002	Protein, Crude (%)	2059	32.18	0.2906	31.94	0.3827	0.1856	187	0.64	002.06	0
002	Protein, Crude (%)	0512	32.19	0.5300	31.94	0.3827	0.1856	187	0.65	002.06	0
002	Protein, Crude (%)	0589	32.19	0.1200	31.94	0.3827	0.1856	187	0.66	002.06	0
002	Protein, Crude (%)	2089	32.19	0.0200	31.94	0.3827	0.1856	187	0.66	002.06	0
002	Protein, Crude (%)	0036	32.20	0.0200	31.94	0.3827	0.1856	187	0.68	002.06	0
002	Protein, Crude (%)	0042	32.20	0.0000	31.94	0.3827	0.1856	187	0.68	002.06	0
002	Protein, Crude (%)	0910	32.20	0.6000	31.94	0.3827	0.1856	187	0.68	002.06	0
002	Protein, Crude (%)	0407	32.20	0.0130	31.94	0.3827	0.1856	187	0.69	002.06	0
002	Protein, Crude (%)	0511	32.21	0.2400	31.94	0.3827	0.1856	187	0.71	002.06	0
002	Protein, Crude (%)	0590	32.25	0.1000	31.94	0.3827	0.1856	187	0.82	002.06	0
002	Protein, Crude (%)	0425	32.28	0.0400	31.94	0.3827	0.1856	187	0.89	002.06	0
002	Protein, Crude (%)	0553	32.29	0.6500	31.94	0.3827	0.1856	187	0.91	002.06	0
002	Protein, Crude (%)	2150	32.31	0.2800	31.94	0.3827	0.1856	187	0.97	002.06	0
002	Protein, Crude (%)	2081	32.32	0.2400	31.94	0.3827	0.1856	187	1.00	002.06	0
002	Protein, Crude (%)	0190	32.33	0.0300	31.94	0.3827	0.1856	187	1.01	002.06	0
002	Protein, Crude (%)	0880	32.35	0.3000	31.94	0.3827	0.1856	187	1.08	002.06	0
002	Protein, Crude (%)	0571	32.37	0.0800	31.94	0.3827	0.1856	187	1.13	002.06	0
002	Protein, Crude (%)	2012	32.37	0.2200	31.94	0.3827	0.1856	187	1.13	002.06	0
002	Protein, Crude (%)	2242	32.38	0.4300	31.94	0.3827	0.1856	187	1.14	002.06	0
002	Protein, Crude (%)	0148	32.38	0.0010	31.94	0.3827	0.1856	187	1.16	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	2307	32.39	0.0300	31.94	0.3827	0.1856	187	1.17	002.06	0
002	Protein, Crude (%)	0089	32.39	0.0000	31.94	0.3827	0.1856	187	1.18	002.06	0
002	Protein, Crude (%)	0037	32.40	0.1500	31.94	0.3827	0.1856	187	1.19	002.06	0
002	Protein, Crude (%)	0357	32.41	0.0625	31.94	0.3827	0.1856	187	1.22	002.06	0
002	Protein, Crude (%)	0186	32.42	0.0500	31.94	0.3827	0.1856	187	1.25	002.06	0
002	Protein, Crude (%)	0643	32.44	0.1400	31.94	0.3827	0.1856	187	1.31	002.99	0
002	Protein, Crude (%)	0045	32.45	0.3000	31.94	0.3827	0.1856	187	1.34	002.06	0
002	Protein, Crude (%)	0001	32.46	0.7200	31.94	0.3827	0.1856	187	1.36	002.06	0
002	Protein, Crude (%)	0687	32.48	0.4500	31.94	0.3827	0.1856	187	1.40	002.06	0
002	Protein, Crude (%)	2076	32.48	0.0220	31.94	0.3827	0.1856	187	1.41	002.06	0
002	Protein, Crude (%)	0242	32.48	0.2800	31.94	0.3827	0.1856	187	1.42	002.06	0
002	Protein, Crude (%)	0083	32.58	0.1800	31.94	0.3827	0.1856	187	1.68	002.06	0
002	Protein, Crude (%)	0843	32.61	0.1000	31.94	0.3827	0.1856	187	1.76	002.06	0
002	Protein, Crude (%)	0300	32.63	0.8900	31.94	0.3827	0.1856	187	1.79	002.06	0
002	Protein, Crude (%)	0164	32.66	0.3100	31.94	0.3827	0.1856	187	1.87	002.06	0
002	Protein, Crude (%)	0417	32.72	0.6100	31.94	0.3827	0.1856	187	2.03	002.06	0
002	Protein, Crude (%)	2213	32.75	0.0100	31.94	0.3827	0.1856	187	2.11	002.06	0
002	Protein, Crude (%)	0948	32.76	0.1700	31.94	0.3827	0.1856	187	2.13	002.06	0
002	Protein, Crude (%)	2155	32.77	0.3300	31.94	0.3827	0.1856	187	2.16	002.06	0
002	Protein, Crude (%)	0011	32.90	0.4000	31.94	0.3827	0.1856	187	2.51	002.06	0
002	Protein, Crude (%)	0685	33.05	0.0200	31.94	0.3827	0.1856	187	2.91	002.01	0
002	Protein, Crude (%)	0618	33.11	0.2220	31.94	0.3827	0.1856	187	3.05	002.06	0
002	Protein, Crude (%)	0405	33.25	0.0300	31.94	0.3827	0.1856	187	3.41	002.11	0
002	Protein, Crude (%)	0889	33.26	0.1900	31.94	0.3827	0.1856	187	3.44	002.11	0
002	Protein, Crude (%)	0011	33.30	0.0000	31.94	0.3827	0.1856	187	3.56	002.11	0
002	Protein, Crude (%)	0033	32.00	1.200	31.94	0.3827	0.1856	187	0.16	002.06	1
002	Protein, Crude (%)	0539	32.03	1.060	31.94	0.3827	0.1856	187	0.24	002.06	1
002	Protein, Crude (%)	0358	32.54	0.9300	31.94	0.3827	0.1856	187	1.56	002.06	1
002	Protein, Crude (%)	0889	0.4600	0.0000	31.94	0.3827	0.1856	187	-82.24	002.01	2
002	Protein, Crude (%)	0620	11.19	0.0000	31.94	0.3827	0.1856	187	-54.21	002.05	2
003	Fat, Crude (%)	2089	6.855	0.1300	7.592	0.2064	0.1113	119	-3.57	003.13	0
003	Fat, Crude (%)	0226	6.885	0.0100	7.592	0.2064	0.1113	119	-3.42	003.09	0
003	Fat, Crude (%)	2295	7.055	0.0100	7.592	0.2064	0.1113	119	-2.60	003.12	0
003	Fat, Crude (%)	2207	7.120	0.1000	7.592	0.2064	0.1113	119	-2.29	003.14	0
003	Fat, Crude (%)	0037	7.170	0.0400	7.592	0.2064	0.1113	119	-2.04	003.14	0
003	Fat, Crude (%)	0646	7.260	0.4800	7.592	0.2064	0.1113	119	-1.61	003.13	0
003	Fat, Crude (%)	0242	7.265	0.0300	7.592	0.2064	0.1113	119	-1.58	003.10	0
003	Fat, Crude (%)	0884	7.300	0.6000	7.592	0.2064	0.1113	119	-1.41	003.06	0
003	Fat, Crude (%)	2162	7.305	0.0900	7.592	0.2064	0.1113	119	-1.39	003.10	0
003	Fat, Crude (%)	2129	7.313	0.0610	7.592	0.2064	0.1113	119	-1.35	003.14	0
003	Fat, Crude (%)	0353	7.320	0.0600	7.592	0.2064	0.1113	119	-1.32	003.09	0
003	Fat, Crude (%)	0682	7.340	0.0000	7.592	0.2064	0.1113	119	-1.22	003.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
003	Fat, Crude (%)	0889	7.340	0.0300	7.592	0.2064	0.1113	119	-1.22	003.11	0
003	Fat, Crude (%)	0512	7.358	0.5210	7.592	0.2064	0.1113	119	-1.14	003.09	0
003	Fat, Crude (%)	0298	7.370	0.0200	7.592	0.2064	0.1113	119	-1.07	003.14	0
003	Fat, Crude (%)	0536	7.380	0.1200	7.592	0.2064	0.1113	119	-1.03	003.99	0
003	Fat, Crude (%)	2181	7.390	0.1200	7.592	0.2064	0.1113	119	-0.98	003.06	0
003	Fat, Crude (%)	0870	7.393	0.0476	7.592	0.2064	0.1113	119	-0.97	003.10	0
003	Fat, Crude (%)	0693	7.397	0.0130	7.592	0.2064	0.1113	119	-0.95	003.10	0
003	Fat, Crude (%)	2076	7.397	0.3360	7.592	0.2064	0.1113	119	-0.94	003.14	0
003	Fat, Crude (%)	0144	7.415	0.0100	7.592	0.2064	0.1113	119	-0.86	003.14	0
003	Fat, Crude (%)	0675	7.420	0.1800	7.592	0.2064	0.1113	119	-0.83	003.09	0
003	Fat, Crude (%)	0948	7.420	0.0800	7.592	0.2064	0.1113	119	-0.83	003.09	0
003	Fat, Crude (%)	0036	7.427	0.0874	7.592	0.2064	0.1113	119	-0.80	003.10	0
003	Fat, Crude (%)	2188	7.430	0.0400	7.592	0.2064	0.1113	119	-0.78	003.10	0
003	Fat, Crude (%)	0045	7.435	0.0700	7.592	0.2064	0.1113	119	-0.76	003.10	0
003	Fat, Crude (%)	0870	7.443	0.3035	7.592	0.2064	0.1113	119	-0.72	003.09	0
003	Fat, Crude (%)	0687	7.450	0.3000	7.592	0.2064	0.1113	119	-0.69	003.06	0
003	Fat, Crude (%)	0638	7.450	0.2000	7.592	0.2064	0.1113	119	-0.69	003.09	0
003	Fat, Crude (%)	2155	7.460	0.0890	7.592	0.2064	0.1113	119	-0.64	003.10	0
003	Fat, Crude (%)	0644	7.467	0.0610	7.592	0.2064	0.1113	119	-0.61	003.06	0
003	Fat, Crude (%)	0042	7.470	0.3800	7.592	0.2064	0.1113	119	-0.59	003.10	0
003	Fat, Crude (%)	0066	7.475	0.0900	7.592	0.2064	0.1113	119	-0.57	003.14	0
003	Fat, Crude (%)	2136	7.480	0.0200	7.592	0.2064	0.1113	119	-0.54	003.10	0
003	Fat, Crude (%)	0186	7.480	0.0400	7.592	0.2064	0.1113	119	-0.54	003.14	0
003	Fat, Crude (%)	2115	7.485	0.4100	7.592	0.2064	0.1113	119	-0.52	003.14	0
003	Fat, Crude (%)	2009	7.487	0.0199	7.592	0.2064	0.1113	119	-0.51	003.10	0
003	Fat, Crude (%)	0202	7.490	0.3000	7.592	0.2064	0.1113	119	-0.49	003.14	0
003	Fat, Crude (%)	0529	7.495	0.1700	7.592	0.2064	0.1113	119	-0.47	003.14	0
003	Fat, Crude (%)	0689	7.500	0.0000	7.592	0.2064	0.1113	119	-0.44	003.10	0
003	Fat, Crude (%)	0880	7.500	0.0000	7.592	0.2064	0.1113	119	-0.44	003.99	0
003	Fat, Crude (%)	2006	7.510	0.0400	7.592	0.2064	0.1113	119	-0.40	003.13	0
003	Fat, Crude (%)	0345	7.515	0.1700	7.592	0.2064	0.1113	119	-0.37	003.00	0
003	Fat, Crude (%)	2109	7.515	0.0500	7.592	0.2064	0.1113	119	-0.37	003.00	0
003	Fat, Crude (%)	0297	7.515	0.0300	7.592	0.2064	0.1113	119	-0.37	003.06	0
003	Fat, Crude (%)	2073	7.515	0.0700	7.592	0.2064	0.1113	119	-0.37	003.10	0
003	Fat, Crude (%)	0889	7.530	0.4800	7.592	0.2064	0.1113	119	-0.30	003.10	0
003	Fat, Crude (%)	0175	7.535	0.0700	7.592	0.2064	0.1113	119	-0.28	003.14	0
003	Fat, Crude (%)	0358	7.540	0.0400	7.592	0.2064	0.1113	119	-0.25	003.14	0
003	Fat, Crude (%)	0541	7.545	0.0900	7.592	0.2064	0.1113	119	-0.23	003.10	0
003	Fat, Crude (%)	0294	7.550	0.0200	7.592	0.2064	0.1113	119	-0.20	003.06	0
003	Fat, Crude (%)	0674	7.550	0.0000	7.592	0.2064	0.1113	119	-0.20	003.09	0
003	Fat, Crude (%)	0098	7.550	0.1600	7.592	0.2064	0.1113	119	-0.20	003.10	0
003	Fat, Crude (%)	0089	7.550	0.0000	7.592	0.2064	0.1113	119	-0.20	003.14	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
003	Fat, Crude (%)	0956	7.550	0.1000	7.592	0.2064	0.1113	119	-0.20	003.14	0
003	Fat, Crude (%)	0164	7.555	0.0300	7.592	0.2064	0.1113	119	-0.18	003.01	0
003	Fat, Crude (%)	0897	7.560	0.0000	7.592	0.2064	0.1113	119	-0.15	003.10	0
003	Fat, Crude (%)	0939	7.575	0.0100	7.592	0.2064	0.1113	119	-0.08	003.06	0
003	Fat, Crude (%)	0683	7.580	0.1200	7.592	0.2064	0.1113	119	-0.06	003.00	0
003	Fat, Crude (%)	0309	7.583	0.0626	7.592	0.2064	0.1113	119	-0.04	003.00	0
003	Fat, Crude (%)	2192	7.585	0.1100	7.592	0.2064	0.1113	119	-0.03	003.00	0
003	Fat, Crude (%)	0574	7.585	0.0500	7.592	0.2064	0.1113	119	-0.03	003.14	0
003	Fat, Crude (%)	0581	7.585	0.0100	7.592	0.2064	0.1113	119	-0.03	003.14	0
003	Fat, Crude (%)	0001	7.593	0.0200	7.592	0.2064	0.1113	119	0.01	003.14	0
003	Fat, Crude (%)	0878	7.595	0.0500	7.592	0.2064	0.1113	119	0.02	003.14	0
003	Fat, Crude (%)	0366	7.600	0.2000	7.592	0.2064	0.1113	119	0.04	003.10	0
003	Fat, Crude (%)	0726	7.610	0.0400	7.592	0.2064	0.1113	119	0.09	003.14	0
003	Fat, Crude (%)	2193	7.615	0.0500	7.592	0.2064	0.1113	119	0.11	003.10	0
003	Fat, Crude (%)	0619	7.620	0.0400	7.592	0.2064	0.1113	119	0.14	003.10	0
003	Fat, Crude (%)	2152	7.620	0.0800	7.592	0.2064	0.1113	119	0.14	003.12	0
003	Fat, Crude (%)	0175	7.625	0.0300	7.592	0.2064	0.1113	119	0.16	003.00	0
003	Fat, Crude (%)	0148	7.625	0.2100	7.592	0.2064	0.1113	119	0.16	003.14	0
003	Fat, Crude (%)	0598	7.625	0.0100	7.592	0.2064	0.1113	119	0.16	003.14	0
003	Fat, Crude (%)	0100	7.640	0.1600	7.592	0.2064	0.1113	119	0.23	003.10	0
003	Fat, Crude (%)	0619	7.645	0.0700	7.592	0.2064	0.1113	119	0.26	003.14	0
003	Fat, Crude (%)	0123	7.650	0.1200	7.592	0.2064	0.1113	119	0.28	003.13	0
003	Fat, Crude (%)	0876	7.650	0.1000	7.592	0.2064	0.1113	119	0.28	003.14	0
003	Fat, Crude (%)	2196	7.660	0.0000	7.592	0.2064	0.1113	119	0.33	003.10	0
003	Fat, Crude (%)	0910	7.680	0.0200	7.592	0.2064	0.1113	119	0.43	003.06	0
003	Fat, Crude (%)	0148	7.680	0.0800	7.592	0.2064	0.1113	119	0.43	003.10	0
003	Fat, Crude (%)	0019	7.680	0.0000	7.592	0.2064	0.1113	119	0.43	003.14	0
003	Fat, Crude (%)	0968	7.691	0.0790	7.592	0.2064	0.1113	119	0.48	003.14	0
003	Fat, Crude (%)	0868	7.695	0.0700	7.592	0.2064	0.1113	119	0.50	003.14	0
003	Fat, Crude (%)	0689	7.700	0.4000	7.592	0.2064	0.1113	119	0.52	003.06	0
003	Fat, Crude (%)	0685	7.700	0.0200	7.592	0.2064	0.1113	119	0.52	003.09	0
003	Fat, Crude (%)	0692	7.700	0.4000	7.592	0.2064	0.1113	119	0.52	003.14	0
003	Fat, Crude (%)	0169	7.710	0.0800	7.592	0.2064	0.1113	119	0.57	003.06	0
003	Fat, Crude (%)	0098	7.720	0.1800	7.592	0.2064	0.1113	119	0.62	003.13	0
003	Fat, Crude (%)	0098	7.720	0.2800	7.592	0.2064	0.1113	119	0.62	003.09	0
003	Fat, Crude (%)	2022	7.725	0.1700	7.592	0.2064	0.1113	119	0.65	003.10	0
003	Fat, Crude (%)	0618	7.726	0.0250	7.592	0.2064	0.1113	119	0.65	003.10	0
003	Fat, Crude (%)	0083	7.730	0.1200	7.592	0.2064	0.1113	119	0.67	003.14	0
003	Fat, Crude (%)	0142	7.740	0.1200	7.592	0.2064	0.1113	119	0.72	003.00	0
003	Fat, Crude (%)	0278	7.750	0.1000	7.592	0.2064	0.1113	119	0.77	003.14	0
003	Fat, Crude (%)	0357	7.750	0.1000	7.592	0.2064	0.1113	119	0.77	003.14	0
003	Fat, Crude (%)	0265	7.760	0.1400	7.592	0.2064	0.1113	119	0.81	003.14	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
003	Fat, Crude (%)	0948	7.760	0.1600	7.592	0.2064	0.1113	119	0.81	003.14	0
003	Fat, Crude (%)	0407	7.784	0.0378	7.592	0.2064	0.1113	119	0.93	003.14	0
003	Fat, Crude (%)	2291	7.785	0.0300	7.592	0.2064	0.1113	119	0.94	003.13	0
003	Fat, Crude (%)	2033	7.790	0.0400	7.592	0.2064	0.1113	119	0.96	003.06	0
003	Fat, Crude (%)	2012	7.805	0.0500	7.592	0.2064	0.1113	119	1.03	003.09	0
003	Fat, Crude (%)	0918	7.825	0.0300	7.592	0.2064	0.1113	119	1.13	003.06	0
003	Fat, Crude (%)	0354	7.825	0.0900	7.592	0.2064	0.1113	119	1.13	003.09	0
003	Fat, Crude (%)	0062	7.835	0.1460	7.592	0.2064	0.1113	119	1.18	003.10	0
003	Fat, Crude (%)	0190	7.855	0.0500	7.592	0.2064	0.1113	119	1.28	003.14	0
003	Fat, Crude (%)	0682	7.880	0.0000	7.592	0.2064	0.1113	119	1.40	003.14	0
003	Fat, Crude (%)	2115	7.915	0.0900	7.592	0.2064	0.1113	119	1.57	003.00	0
003	Fat, Crude (%)	0870	7.989	0.2993	7.592	0.2064	0.1113	119	1.93	003.13	0
003	Fat, Crude (%)	0590	8.235	0.0700	7.592	0.2064	0.1113	119	3.12	003.09	0
003	Fat, Crude (%)	0520	8.245	0.0100	7.592	0.2064	0.1113	119	3.16	003.14	0
003	Fat, Crude (%)	0011	9.300	0.0000	7.592	0.2064	0.1113	119	8.28	003.11	0
003	Fat, Crude (%)	0297	9.755	0.0700	7.592	0.2064	0.1113	119	10.48	003.11	0
003	Fat, Crude (%)	2232	9.805	0.0300	7.592	0.2064	0.1113	119	10.72	003.11	0
003	Fat, Crude (%)	0860	10.10	0.0000	7.592	0.2064	0.1113	119	12.15	003.14	0
003	Fat, Crude (%)	0853	10.20	0.2000	7.592	0.2064	0.1113	119	12.64	003.14	0
003	Fat, Crude (%)	0405	10.56	0.0100	7.592	0.2064	0.1113	119	14.36	003.11	0
003	Fat, Crude (%)	0941	10.60	0.1400	7.592	0.2064	0.1113	119	14.57	003.99	0
003	Fat, Crude (%)	0009	10.69	0.5500	7.592	0.2064	0.1113	119	14.99	003.14	0
003	Fat, Crude (%)	0033	11.10	0.2000	7.592	0.2064	0.1113	119	17.00	003.13	0
003	Fat, Crude (%)	0300	7.300	0.6200	7.592	0.2064	0.1113	119	-1.41	003.14	1
003	Fat, Crude (%)	2268	10.54	0.6300	7.592	0.2064	0.1113	119	14.26	003.06	1
003	Fat, Crude (%)	0038	11.37	0.6800	7.592	0.2064	0.1113	119	18.28	003.09	1
004	Fiber, Crude (%)	0042	1.925	0.0900	4.657	0.6849	0.2001	109	-3.99	004.07	0
004	Fiber, Crude (%)	2089	2.330	0.2400	4.657	0.6849	0.2001	109	-3.40	004.03	0
004	Fiber, Crude (%)	0202	2.900	1.160	4.657	0.6849	0.2001	109	-2.57	004.07	0
004	Fiber, Crude (%)	0405	2.915	0.0100	4.657	0.6849	0.2001	109	-2.54	004.11	0
004	Fiber, Crude (%)	2304	3.445	0.2700	4.657	0.6849	0.2001	109	-1.77	004.07	0
004	Fiber, Crude (%)	0675	3.540	0.1400	4.657	0.6849	0.2001	109	-1.63	004.07	0
004	Fiber, Crude (%)	0190	3.550	0.1000	4.657	0.6849	0.2001	109	-1.62	004.07	0
004	Fiber, Crude (%)	0708	3.600	0.2000	4.657	0.6849	0.2001	109	-1.54	004.07	0
004	Fiber, Crude (%)	0956	3.600	0.0000	4.657	0.6849	0.2001	109	-1.54	004.07	0
004	Fiber, Crude (%)	0921	3.645	0.4700	4.657	0.6849	0.2001	109	-1.48	004.03	0
004	Fiber, Crude (%)	2115	3.720	0.0800	4.657	0.6849	0.2001	109	-1.37	004.07	0
004	Fiber, Crude (%)	2053	3.805	0.0900	4.657	0.6849	0.2001	109	-1.24	004.07	0
004	Fiber, Crude (%)	2146	3.875	0.2300	4.657	0.6849	0.2001	109	-1.14	004.99	0
004	Fiber, Crude (%)	0868	3.885	0.0300	4.657	0.6849	0.2001	109	-1.13	004.06	0
004	Fiber, Crude (%)	2109	3.895	0.1700	4.657	0.6849	0.2001	109	-1.11	004.07	0
004	Fiber, Crude (%)	0682	3.900	0.0000	4.657	0.6849	0.2001	109	-1.11	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
004	Fiber, Crude (%)	2188	3.910	0.0000	4.657	0.6849	0.2001	109	-1.09	004.07	0
004	Fiber, Crude (%)	0590	3.915	0.0100	4.657	0.6849	0.2001	109	-1.08	004.07	0
004	Fiber, Crude (%)	0169	4.100	0.2000	4.657	0.6849	0.2001	109	-0.81	004.00	0
004	Fiber, Crude (%)	0407	4.128	0.1541	4.657	0.6849	0.2001	109	-0.77	004.07	0
004	Fiber, Crude (%)	0644	4.133	0.1040	4.657	0.6849	0.2001	109	-0.77	004.07	0
004	Fiber, Crude (%)	0175	4.140	0.1200	4.657	0.6849	0.2001	109	-0.76	004.00	0
004	Fiber, Crude (%)	0848	4.145	0.1100	4.657	0.6849	0.2001	109	-0.75	004.07	0
004	Fiber, Crude (%)	0019	4.185	0.3700	4.657	0.6849	0.2001	109	-0.69	004.07	0
004	Fiber, Crude (%)	0726	4.260	0.0000	4.657	0.6849	0.2001	109	-0.58	004.07	0
004	Fiber, Crude (%)	0939	4.285	0.0300	4.657	0.6849	0.2001	109	-0.54	004.07	0
004	Fiber, Crude (%)	0098	4.305	0.1900	4.657	0.6849	0.2001	109	-0.51	004.06	0
004	Fiber, Crude (%)	0968	4.325	0.2430	4.657	0.6849	0.2001	109	-0.49	004.06	0
004	Fiber, Crude (%)	0425	4.350	0.1000	4.657	0.6849	0.2001	109	-0.45	004.00	0
004	Fiber, Crude (%)	0300	4.365	0.0100	4.657	0.6849	0.2001	109	-0.43	004.07	0
004	Fiber, Crude (%)	0265	4.375	0.7500	4.657	0.6849	0.2001	109	-0.41	004.07	0
004	Fiber, Crude (%)	0968	4.396	0.0680	4.657	0.6849	0.2001	109	-0.38	004.07	0
004	Fiber, Crude (%)	2136	4.400	0.7800	4.657	0.6849	0.2001	109	-0.38	004.07	0
004	Fiber, Crude (%)	0144	4.410	0.2200	4.657	0.6849	0.2001	109	-0.36	004.07	0
004	Fiber, Crude (%)	0860	4.440	0.0000	4.657	0.6849	0.2001	109	-0.32	004.07	0
004	Fiber, Crude (%)	2115	4.445	0.0100	4.657	0.6849	0.2001	109	-0.31	004.06	0
004	Fiber, Crude (%)	2009	4.445	0.0389	4.657	0.6849	0.2001	109	-0.31	004.07	0
004	Fiber, Crude (%)	0876	4.450	0.9000	4.657	0.6849	0.2001	109	-0.30	004.07	0
004	Fiber, Crude (%)	2207	4.480	0.0400	4.657	0.6849	0.2001	109	-0.26	004.07	0
004	Fiber, Crude (%)	0242	4.485	0.1700	4.657	0.6849	0.2001	109	-0.25	004.07	0
004	Fiber, Crude (%)	0683	4.490	0.3000	4.657	0.6849	0.2001	109	-0.24	004.07	0
004	Fiber, Crude (%)	0083	4.495	0.0100	4.657	0.6849	0.2001	109	-0.24	004.07	0
004	Fiber, Crude (%)	0619	4.495	0.1100	4.657	0.6849	0.2001	109	-0.24	004.07	0
004	Fiber, Crude (%)	0164	4.500	0.0000	4.657	0.6849	0.2001	109	-0.23	004.00	0
004	Fiber, Crude (%)	0910	4.500	0.2000	4.657	0.6849	0.2001	109	-0.23	004.07	0
004	Fiber, Crude (%)	2033	4.500	0.0000	4.657	0.6849	0.2001	109	-0.23	004.07	0
004	Fiber, Crude (%)	0610	4.510	0.0400	4.657	0.6849	0.2001	109	-0.21	004.99	0
004	Fiber, Crude (%)	0354	4.530	0.0200	4.657	0.6849	0.2001	109	-0.19	004.07	0
004	Fiber, Crude (%)	0674	4.550	0.0000	4.657	0.6849	0.2001	109	-0.16	004.06	0
004	Fiber, Crude (%)	0510	4.550	0.1000	4.657	0.6849	0.2001	109	-0.16	004.07	0
004	Fiber, Crude (%)	0529	4.560	0.0000	4.657	0.6849	0.2001	109	-0.14	004.07	0
004	Fiber, Crude (%)	2181	4.560	0.0000	4.657	0.6849	0.2001	109	-0.14	004.07	0
004	Fiber, Crude (%)	2022	4.560	0.2000	4.657	0.6849	0.2001	109	-0.14	004.06	0
004	Fiber, Crude (%)	0100	4.580	0.7200	4.657	0.6849	0.2001	109	-0.11	004.07	0
004	Fiber, Crude (%)	0653	4.590	0.2600	4.657	0.6849	0.2001	109	-0.10	004.07	0
004	Fiber, Crude (%)	0009	4.615	0.3700	4.657	0.6849	0.2001	109	-0.06	004.07	0
004	Fiber, Crude (%)	0098	4.615	0.0300	4.657	0.6849	0.2001	109	-0.06	004.07	0
004	Fiber, Crude (%)	0512	4.619	0.2940	4.657	0.6849	0.2001	109	-0.06	004.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
004	Fiber, Crude (%)	0353	4.630	0.0800	4.657	0.6849	0.2001	109	-0.04	004.03	0
004	Fiber, Crude (%)	0263	4.635	0.0690	4.657	0.6849	0.2001	109	-0.03	004.07	0
004	Fiber, Crude (%)	0870	4.638	0.2423	4.657	0.6849	0.2001	109	-0.03	004.07	0
004	Fiber, Crude (%)	2193	4.655	0.0700	4.657	0.6849	0.2001	109	0.00	004.07	0
004	Fiber, Crude (%)	0066	4.700	0.2600	4.657	0.6849	0.2001	109	0.06	004.07	0
004	Fiber, Crude (%)	0536	4.700	0.0000	4.657	0.6849	0.2001	109	0.06	004.07	0
004	Fiber, Crude (%)	0948	4.714	0.0550	4.657	0.6849	0.2001	109	0.08	004.07	0
004	Fiber, Crude (%)	0948	4.719	0.1030	4.657	0.6849	0.2001	109	0.09	004.06	0
004	Fiber, Crude (%)	2006	4.725	0.0100	4.657	0.6849	0.2001	109	0.10	004.06	0
004	Fiber, Crude (%)	0123	4.740	0.0800	4.657	0.6849	0.2001	109	0.12	004.07	0
004	Fiber, Crude (%)	2307	4.740	0.0200	4.657	0.6849	0.2001	109	0.12	004.07	0
004	Fiber, Crude (%)	0598	4.780	0.0400	4.657	0.6849	0.2001	109	0.18	004.07	0
004	Fiber, Crude (%)	0878	4.810	0.0800	4.657	0.6849	0.2001	109	0.22	004.06	0
004	Fiber, Crude (%)	0354	4.815	0.1500	4.657	0.6849	0.2001	109	0.23	004.00	0
004	Fiber, Crude (%)	0345	4.825	0.1500	4.657	0.6849	0.2001	109	0.24	004.00	0
004	Fiber, Crude (%)	0309	4.845	0.0580	4.657	0.6849	0.2001	109	0.27	004.00	0
004	Fiber, Crude (%)	2196	4.850	0.0000	4.657	0.6849	0.2001	109	0.28	004.07	0
004	Fiber, Crude (%)	0581	4.900	0.2000	4.657	0.6849	0.2001	109	0.35	004.07	0
004	Fiber, Crude (%)	0297	4.910	0.6400	4.657	0.6849	0.2001	109	0.37	004.07	0
004	Fiber, Crude (%)	2295	4.965	0.0100	4.657	0.6849	0.2001	109	0.45	004.06	0
004	Fiber, Crude (%)	0553	4.965	0.3900	4.657	0.6849	0.2001	109	0.45	004.07	0
004	Fiber, Crude (%)	2129	5.012	0.0810	4.657	0.6849	0.2001	109	0.52	004.99	0
004	Fiber, Crude (%)	2073	5.050	0.0400	4.657	0.6849	0.2001	109	0.57	004.00	0
004	Fiber, Crude (%)	0941	5.075	0.2100	4.657	0.6849	0.2001	109	0.61	004.06	0
004	Fiber, Crude (%)	2076	5.148	0.5782	4.657	0.6849	0.2001	109	0.72	004.00	0
004	Fiber, Crude (%)	2192	5.150	0.0200	4.657	0.6849	0.2001	109	0.72	004.03	0
004	Fiber, Crude (%)	0689	5.150	0.5000	4.657	0.6849	0.2001	109	0.72	004.06	0
004	Fiber, Crude (%)	0294	5.150	0.1000	4.657	0.6849	0.2001	109	0.72	004.07	0
004	Fiber, Crude (%)	0511	5.250	0.7000	4.657	0.6849	0.2001	109	0.87	004.07	0
004	Fiber, Crude (%)	2264	5.288	0.6008	4.657	0.6849	0.2001	109	0.92	004.00	0
004	Fiber, Crude (%)	0638	5.300	0.6000	4.657	0.6849	0.2001	109	0.94	004.06	0
004	Fiber, Crude (%)	0897	5.320	0.0000	4.657	0.6849	0.2001	109	0.97	004.06	0
004	Fiber, Crude (%)	0918	5.370	0.2200	4.657	0.6849	0.2001	109	1.04	004.06	0
004	Fiber, Crude (%)	0226	5.395	0.4500	4.657	0.6849	0.2001	109	1.08	004.00	0
004	Fiber, Crude (%)	0045	5.400	1.200	4.657	0.6849	0.2001	109	1.08	004.07	0
004	Fiber, Crude (%)	0298	5.400	0.2000	4.657	0.6849	0.2001	109	1.08	004.07	0
004	Fiber, Crude (%)	0845	5.515	0.3500	4.657	0.6849	0.2001	109	1.25	004.06	0
004	Fiber, Crude (%)	0904	5.550	0.3200	4.657	0.6849	0.2001	109	1.30	004.03	0
004	Fiber, Crude (%)	2291	5.555	0.2100	4.657	0.6849	0.2001	109	1.31	004.06	0
004	Fiber, Crude (%)	2314	5.750	0.1000	4.657	0.6849	0.2001	109	1.60	004.06	0
004	Fiber, Crude (%)	0520	5.800	0.2000	4.657	0.6849	0.2001	109	1.67	004.07	0
004	Fiber, Crude (%)	0034	5.820	0.2000	4.657	0.6849	0.2001	109	1.70	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
004	Fiber, Crude (%)	2259	5.828	0.0250	4.657	0.6849	0.2001	109	1.71	004.07	0
004	Fiber, Crude (%)	0884	5.850	0.3000	4.657	0.6849	0.2001	109	1.74	004.00	0
004	Fiber, Crude (%)	0089	5.915	0.0300	4.657	0.6849	0.2001	109	1.84	004.07	0
004	Fiber, Crude (%)	0554	5.985	0.0500	4.657	0.6849	0.2001	109	1.94	004.07	0
004	Fiber, Crude (%)	2012	6.475	0.1300	4.657	0.6849	0.2001	109	2.65	004.07	0
004	Fiber, Crude (%)	0011	6.900	0.0000	4.657	0.6849	0.2001	109	3.27	004.11	0
004	Fiber, Crude (%)	0889	6.928	0.0650	4.657	0.6849	0.2001	109	3.31	004.11	0
004	Fiber, Crude (%)	0366	6.950	0.5000	4.657	0.6849	0.2001	109	3.35	004.01	0
004	Fiber, Crude (%)	0693	7.093	0.5750	4.657	0.6849	0.2001	109	3.56	004.07	0
004	Fiber, Crude (%)	0017	5.350	1.300	4.657	0.6849	0.2001	109	1.01	004.00	1
004	Fiber, Crude (%)	0643	6.845	2.270	4.657	0.6849	0.2001	109	3.19	004.07	1
004	Fiber, Crude (%)	0278	11.20	0.2000	4.657	0.6849	0.2001	109	9.55	004.07	2
004	Fiber, Crude (%)	0889	0.0000	0.0000	4.657	0.6849	0.2001	109		004.07	4
005	Ash (%)	0674	6.160	0.0000	6.885	0.1167	0.0710	127	-6.21	005.00	0
005	Ash (%)	2081	6.180	0.0600	6.885	0.1167	0.0710	127	-6.04	005.99	0
005	Ash (%)	2169	6.385	0.0500	6.885	0.1167	0.0710	127	-4.28	005.99	0
005	Ash (%)	0638	6.450	0.0200	6.885	0.1167	0.0710	127	-3.73	005.05	0
005	Ash (%)	0644	6.515	0.0700	6.885	0.1167	0.0710	127	-3.17	005.02	0
005	Ash (%)	0589	6.540	0.2000	6.885	0.1167	0.0710	127	-2.95	005.00	0
005	Ash (%)	0169	6.600	0.0200	6.885	0.1167	0.0710	127	-2.44	005.00	0
005	Ash (%)	0425	6.600	0.2000	6.885	0.1167	0.0710	127	-2.44	005.00	0
005	Ash (%)	2033	6.650	0.1000	6.885	0.1167	0.0710	127	-2.01	005.00	0
005	Ash (%)	0358	6.665	0.1500	6.885	0.1167	0.0710	127	-1.88	005.00	0
005	Ash (%)	0019	6.700	0.0000	6.885	0.1167	0.0710	127	-1.58	005.00	0
005	Ash (%)	0265	6.700	0.0000	6.885	0.1167	0.0710	127	-1.58	005.00	0
005	Ash (%)	0294	6.700	0.0000	6.885	0.1167	0.0710	127	-1.58	005.05	0
005	Ash (%)	0918	6.705	0.1300	6.885	0.1167	0.0710	127	-1.54	005.00	0
005	Ash (%)	0042	6.720	0.1200	6.885	0.1167	0.0710	127	-1.41	005.00	0
005	Ash (%)	2129	6.720	0.0660	6.885	0.1167	0.0710	127	-1.41	005.05	0
005	Ash (%)	0880	6.730	0.2200	6.885	0.1167	0.0710	127	-1.33	005.00	0
005	Ash (%)	0853	6.750	0.1000	6.885	0.1167	0.0710	127	-1.16	005.00	0
005	Ash (%)	2307	6.755	0.0100	6.885	0.1167	0.0710	127	-1.11	005.00	0
005	Ash (%)	2268	6.760	0.0600	6.885	0.1167	0.0710	127	-1.07	005.00	0
005	Ash (%)	2259	6.766	0.0350	6.885	0.1167	0.0710	127	-1.02	005.00	0
005	Ash (%)	0574	6.775	0.3300	6.885	0.1167	0.0710	127	-0.94	005.00	0
005	Ash (%)	2022	6.795	0.0300	6.885	0.1167	0.0710	127	-0.77	005.00	0
005	Ash (%)	0618	6.799	0.0279	6.885	0.1167	0.0710	127	-0.73	005.00	0
005	Ash (%)	0366	6.800	0.0000	6.885	0.1167	0.0710	127	-0.73	005.00	0
005	Ash (%)	0692	6.800	0.0000	6.885	0.1167	0.0710	127	-0.73	005.05	0
005	Ash (%)	0278	6.800	0.1400	6.885	0.1167	0.0710	127	-0.73	005.00	0
005	Ash (%)	0598	6.805	0.0100	6.885	0.1167	0.0710	127	-0.68	005.00	0
005	Ash (%)	0539	6.810	0.0600	6.885	0.1167	0.0710	127	-0.64	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
005	Ash (%)	2181	6.810	0.0400	6.885	0.1167	0.0710	127	-0.64	005.00	0
005	Ash (%)	2076	6.811	0.2149	6.885	0.1167	0.0710	127	-0.63	005.00	0
005	Ash (%)	0345	6.815	0.1100	6.885	0.1167	0.0710	127	-0.60	005.00	0
005	Ash (%)	0062	6.817	0.0048	6.885	0.1167	0.0710	127	-0.59	005.00	0
005	Ash (%)	2073	6.825	0.0700	6.885	0.1167	0.0710	127	-0.51	005.00	0
005	Ash (%)	2232	6.825	0.1500	6.885	0.1167	0.0710	127	-0.51	005.00	0
005	Ash (%)	0870	6.828	0.0159	6.885	0.1167	0.0710	127	-0.49	005.00	0
005	Ash (%)	0035	6.830	0.0800	6.885	0.1167	0.0710	127	-0.47	005.00	0
005	Ash (%)	0144	6.830	0.0400	6.885	0.1167	0.0710	127	-0.47	005.00	0
005	Ash (%)	0407	6.830	0.0794	6.885	0.1167	0.0710	127	-0.47	005.00	0
005	Ash (%)	0643	6.830	0.1400	6.885	0.1167	0.0710	127	-0.47	005.00	0
005	Ash (%)	2242	6.830	0.0400	6.885	0.1167	0.0710	127	-0.47	005.00	0
005	Ash (%)	2304	6.830	0.0200	6.885	0.1167	0.0710	127	-0.47	005.03	0
005	Ash (%)	0553	6.835	0.0300	6.885	0.1167	0.0710	127	-0.43	005.00	0
005	Ash (%)	0726	6.835	0.0100	6.885	0.1167	0.0710	127	-0.43	005.05	0
005	Ash (%)	0083	6.840	0.0200	6.885	0.1167	0.0710	127	-0.38	005.00	0
005	Ash (%)	0417	6.840	0.0000	6.885	0.1167	0.0710	127	-0.38	005.00	0
005	Ash (%)	0164	6.845	0.0900	6.885	0.1167	0.0710	127	-0.34	005.00	0
005	Ash (%)	0354	6.845	0.0300	6.885	0.1167	0.0710	127	-0.34	005.00	0
005	Ash (%)	0620	6.847	0.0358	6.885	0.1167	0.0710	127	-0.32	005.00	0
005	Ash (%)	2193	6.850	0.0000	6.885	0.1167	0.0710	127	-0.30	005.00	0
005	Ash (%)	0646	6.855	0.1700	6.885	0.1167	0.0710	127	-0.26	005.00	0
005	Ash (%)	2295	6.855	0.0100	6.885	0.1167	0.0710	127	-0.26	005.05	0
005	Ash (%)	0175	6.870	0.0400	6.885	0.1167	0.0710	127	-0.13	005.00	0
005	Ash (%)	0142	6.870	0.1400	6.885	0.1167	0.0710	127	-0.13	005.00	0
005	Ash (%)	0541	6.870	0.0600	6.885	0.1167	0.0710	127	-0.13	005.00	0
005	Ash (%)	0186	6.875	0.0900	6.885	0.1167	0.0710	127	-0.08	005.00	0
005	Ash (%)	2187	6.875	0.0500	6.885	0.1167	0.0710	127	-0.08	005.03	0
005	Ash (%)	2264	6.884	0.1110	6.885	0.1167	0.0710	127	0.00	005.05	0
005	Ash (%)	0953	6.885	0.2100	6.885	0.1167	0.0710	127	0.00	005.00	0
005	Ash (%)	2312	6.890	0.1200	6.885	0.1167	0.0710	127	0.04	005.00	0
005	Ash (%)	0868	6.890	0.0600	6.885	0.1167	0.0710	127	0.04	005.05	0
005	Ash (%)	2150	6.895	0.0500	6.885	0.1167	0.0710	127	0.09	005.00	0
005	Ash (%)	0123	6.900	0.0400	6.885	0.1167	0.0710	127	0.13	005.00	0
005	Ash (%)	0298	6.900	0.0200	6.885	0.1167	0.0710	127	0.13	005.00	0
005	Ash (%)	0357	6.900	0.0000	6.885	0.1167	0.0710	127	0.13	005.00	0
005	Ash (%)	0876	6.900	0.0000	6.885	0.1167	0.0710	127	0.13	005.00	0
005	Ash (%)	2115	6.900	0.1000	6.885	0.1167	0.0710	127	0.13	005.00	0
005	Ash (%)	0683	6.900	0.1400	6.885	0.1167	0.0710	127	0.13	005.05	0
005	Ash (%)	0845	6.905	0.0500	6.885	0.1167	0.0710	127	0.17	005.00	0
005	Ash (%)	0689	6.905	0.0300	6.885	0.1167	0.0710	127	0.17	005.05	0
005	Ash (%)	2156	6.905	0.0100	6.885	0.1167	0.0710	127	0.17	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
005	Ash (%)	0968	6.908	0.0030	6.885	0.1167	0.0710	127	0.19	005.00	0
005	Ash (%)	0693	6.909	0.0260	6.885	0.1167	0.0710	127	0.21	005.00	0
005	Ash (%)	2196	6.910	0.0000	6.885	0.1167	0.0710	127	0.22	005.00	0
005	Ash (%)	0889	6.910	0.0000	6.885	0.1167	0.0710	127	0.22	005.05	0
005	Ash (%)	2016	6.915	0.1500	6.885	0.1167	0.0710	127	0.26	005.00	0
005	Ash (%)	2109	6.915	0.3500	6.885	0.1167	0.0710	127	0.26	005.05	0
005	Ash (%)	0682	6.920	0.0000	6.885	0.1167	0.0710	127	0.30	005.00	0
005	Ash (%)	2188	6.920	0.0400	6.885	0.1167	0.0710	127	0.30	005.00	0
005	Ash (%)	2124	6.930	0.0800	6.885	0.1167	0.0710	127	0.39	005.00	0
005	Ash (%)	0309	6.932	0.0940	6.885	0.1167	0.0710	127	0.40	005.00	0
005	Ash (%)	0263	6.935	0.0190	6.885	0.1167	0.0710	127	0.43	005.05	0
005	Ash (%)	0148	6.935	0.1300	6.885	0.1167	0.0710	127	0.43	005.00	0
005	Ash (%)	2291	6.935	0.0100	6.885	0.1167	0.0710	127	0.43	005.05	0
005	Ash (%)	0353	6.940	0.0400	6.885	0.1167	0.0710	127	0.47	005.00	0
005	Ash (%)	2006	6.940	0.1400	6.885	0.1167	0.0710	127	0.47	005.00	0
005	Ash (%)	0510	6.940	0.0800	6.885	0.1167	0.0710	127	0.47	005.00	0
005	Ash (%)	2089	6.940	0.0800	6.885	0.1167	0.0710	127	0.47	005.00	0
005	Ash (%)	0139	6.945	0.0500	6.885	0.1167	0.0710	127	0.52	005.00	0
005	Ash (%)	0675	6.945	0.2300	6.885	0.1167	0.0710	127	0.52	005.00	0
005	Ash (%)	0878	6.945	0.0300	6.885	0.1167	0.0710	127	0.52	005.05	0
005	Ash (%)	0941	6.945	0.1700	6.885	0.1167	0.0710	127	0.52	005.05	0
005	Ash (%)	2012	6.945	0.0300	6.885	0.1167	0.0710	127	0.52	005.05	0
005	Ash (%)	0202	6.945	0.1300	6.885	0.1167	0.0710	127	0.52	005.99	0
005	Ash (%)	0956	6.950	0.1000	6.885	0.1167	0.0710	127	0.56	005.00	0
005	Ash (%)	0884	6.950	0.1000	6.885	0.1167	0.0710	127	0.56	005.99	0
005	Ash (%)	0590	6.955	0.0300	6.885	0.1167	0.0710	127	0.60	005.05	0
005	Ash (%)	0619	6.955	0.0900	6.885	0.1167	0.0710	127	0.60	005.05	0
005	Ash (%)	0033	6.965	0.1500	6.885	0.1167	0.0710	127	0.69	005.00	0
005	Ash (%)	0520	6.965	0.0700	6.885	0.1167	0.0710	127	0.69	005.00	0
005	Ash (%)	2009	6.970	0.0361	6.885	0.1167	0.0710	127	0.73	005.05	0
005	Ash (%)	0098	6.975	0.0700	6.885	0.1167	0.0710	127	0.77	005.00	0
005	Ash (%)	2162	6.980	0.0200	6.885	0.1167	0.0710	127	0.82	005.00	0
005	Ash (%)	0610	6.980	0.0600	6.885	0.1167	0.0710	127	0.82	005.05	0
005	Ash (%)	0581	6.985	0.0900	6.885	0.1167	0.0710	127	0.86	005.00	0
005	Ash (%)	2293	6.985	0.0100	6.885	0.1167	0.0710	127	0.86	005.00	0
005	Ash (%)	0045	6.990	0.0200	6.885	0.1167	0.0710	127	0.90	005.00	0
005	Ash (%)	0065	6.993	0.0050	6.885	0.1167	0.0710	127	0.92	005.05	0
005	Ash (%)	2219	6.995	0.0300	6.885	0.1167	0.0710	127	0.94	005.00	0
005	Ash (%)	2220	6.995	0.0100	6.885	0.1167	0.0710	127	0.94	005.00	0
005	Ash (%)	0100	7.020	0.0200	6.885	0.1167	0.0710	127	1.16	005.00	0
005	Ash (%)	0242	7.020	0.1400	6.885	0.1167	0.0710	127	1.16	005.00	0
005	Ash (%)	0653	7.020	0.0000	6.885	0.1167	0.0710	127	1.16	005.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
005	Ash (%)	0897	7.020	0.0000	6.885	0.1167	0.0710	127	1.16	005.05	0
005	Ash (%)	0968	7.029	0.0090	6.885	0.1167	0.0710	127	1.23	005.02	0
005	Ash (%)	2192	7.035	0.1700	6.885	0.1167	0.0710	127	1.29	005.05	0
005	Ash (%)	0536	7.050	0.0000	6.885	0.1167	0.0710	127	1.41	005.99	0
005	Ash (%)	0948	7.055	0.0100	6.885	0.1167	0.0710	127	1.46	005.00	0
005	Ash (%)	0685	7.065	0.0100	6.885	0.1167	0.0710	127	1.54	005.05	0
005	Ash (%)	0652	7.100	0.0000	6.885	0.1167	0.0710	127	1.84	005.99	0
005	Ash (%)	2252	7.105	0.2300	6.885	0.1167	0.0710	127	1.89	005.00	0
005	Ash (%)	0034	7.127	0.1480	6.885	0.1167	0.0710	127	2.07	005.05	0
005	Ash (%)	0676	7.140	0.0600	6.885	0.1167	0.0710	127	2.19	005.99	0
005	Ash (%)	0190	7.155	0.0100	6.885	0.1167	0.0710	127	2.31	005.05	0
005	Ash (%)	0226	7.735	0.1300	6.885	0.1167	0.0710	127	7.28	005.00	0
005	Ash (%)	0297	8.700	0.3000	6.885	0.1167	0.0710	127	15.55	005.11	0
005	Ash (%)	0889	9.200	0.0800	6.885	0.1167	0.0710	127	19.83	005.11	0
005	Ash (%)	0910	7.050	0.5000	6.885	0.1167	0.0710	127	1.41	005.99	1
005	Ash (%)	0297	7.250	0.5000	6.885	0.1167	0.0710	127	3.13	005.00	1
005	Ash (%)	2314	12.85	0.9000	6.885	0.1167	0.0710	127	51.11	005.05	2
006	Total Sugars (%)	0096	1.358	0.0881	2.612	1.057	0.2175	4		006.99	0
006	Total Sugars (%)	0910	2.345	0.0300	2.612	1.057	0.2175	4		006.00	0
006	Total Sugars (%)	0956	2.850	0.5000	2.612	1.057	0.2175	4		006.99	0
006	Total Sugars (%)	0407	3.896	0.2520	2.612	1.057	0.2175	4		006.00	0
008	Fiber, Acid Detergent (%)	0968	3.271	0.1240	6.797	1.215	0.3207	52	-2.90	008.08	0
008	Fiber, Acid Detergent (%)	0590	4.185	0.4300	6.797	1.215	0.3207	52	-2.15	008.08	0
008	Fiber, Acid Detergent (%)	0038	4.487	0.6145	6.797	1.215	0.3207	52	-1.90	008.02	0
008	Fiber, Acid Detergent (%)	0956	4.550	0.1000	6.797	1.215	0.3207	52	-1.85	008.08	0
008	Fiber, Acid Detergent (%)	0042	4.610	0.1800	6.797	1.215	0.3207	52	-1.80	008.08	0
008	Fiber, Acid Detergent (%)	0878	4.835	0.0500	6.797	1.215	0.3207	52	-1.62	008.08	0
008	Fiber, Acid Detergent (%)	0066	5.525	0.1500	6.797	1.215	0.3207	52	-1.05	008.08	0
008	Fiber, Acid Detergent (%)	0653	5.660	0.1000	6.797	1.215	0.3207	52	-0.94	008.08	0
008	Fiber, Acid Detergent (%)	2129	5.790	0.3260	6.797	1.215	0.3207	52	-0.83	008.08	0
008	Fiber, Acid Detergent (%)	0354	5.860	0.0400	6.797	1.215	0.3207	52	-0.77	008.08	0
008	Fiber, Acid Detergent (%)	0675	5.865	0.1500	6.797	1.215	0.3207	52	-0.77	008.08	0
008	Fiber, Acid Detergent (%)	0098	5.945	0.1100	6.797	1.215	0.3207	52	-0.70	008.02	0
008	Fiber, Acid Detergent (%)	0939	6.035	0.0700	6.797	1.215	0.3207	52	-0.63	008.08	0
008	Fiber, Acid Detergent (%)	0083	6.075	0.0500	6.797	1.215	0.3207	52	-0.59	008.08	0
008	Fiber, Acid Detergent (%)	0407	6.164	0.5067	6.797	1.215	0.3207	52	-0.52	008.08	0
008	Fiber, Acid Detergent (%)	0353	6.195	0.1900	6.797	1.215	0.3207	52	-0.50	008.02	0
008	Fiber, Acid Detergent (%)	0309	6.217	0.1701	6.797	1.215	0.3207	52	-0.48	008.02	0
008	Fiber, Acid Detergent (%)	0358	6.305	0.7100	6.797	1.215	0.3207	52	-0.41	008.08	0
008	Fiber, Acid Detergent (%)	0190	6.335	0.0100	6.797	1.215	0.3207	52	-0.38	008.08	0
008	Fiber, Acid Detergent (%)	0848	6.335	0.6900	6.797	1.215	0.3207	52	-0.38	008.08	0
008	Fiber, Acid Detergent (%)	0297	6.570	0.3400	6.797	1.215	0.3207	52	-0.19	008.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
008	Fiber, Acid Detergent (%)	0870	6.678	0.4556	6.797	1.215	0.3207	52	-0.10	008.08	0
008	Fiber, Acid Detergent (%)	0619	6.685	0.0900	6.797	1.215	0.3207	52	-0.09	008.08	0
008	Fiber, Acid Detergent (%)	0948	6.695	0.3100	6.797	1.215	0.3207	52	-0.08	008.02	0
008	Fiber, Acid Detergent (%)	0345	6.700	0.4000	6.797	1.215	0.3207	52	-0.08	008.02	0
008	Fiber, Acid Detergent (%)	0610	6.835	0.0100	6.797	1.215	0.3207	52	0.03	008.99	0
008	Fiber, Acid Detergent (%)	0265	6.890	0.0800	6.797	1.215	0.3207	52	0.08	008.05	0
008	Fiber, Acid Detergent (%)	2136	6.905	1.010	6.797	1.215	0.3207	52	0.09	008.08	0
008	Fiber, Acid Detergent (%)	0910	6.915	0.2900	6.797	1.215	0.3207	52	0.10	008.08	0
008	Fiber, Acid Detergent (%)	0581	7.005	0.2100	6.797	1.215	0.3207	52	0.17	008.08	0
008	Fiber, Acid Detergent (%)	0148	7.010	0.2800	6.797	1.215	0.3207	52	0.18	008.02	0
008	Fiber, Acid Detergent (%)	0263	7.021	0.0600	6.797	1.215	0.3207	52	0.18	008.08	0
008	Fiber, Acid Detergent (%)	0357	7.100	1.000	6.797	1.215	0.3207	52	0.25	008.08	0
008	Fiber, Acid Detergent (%)	2196	7.110	0.0000	6.797	1.215	0.3207	52	0.26	008.08	0
008	Fiber, Acid Detergent (%)	0148	7.135	0.1500	6.797	1.215	0.3207	52	0.28	008.08	0
008	Fiber, Acid Detergent (%)	2192	7.215	0.3700	6.797	1.215	0.3207	52	0.34	008.02	0
008	Fiber, Acid Detergent (%)	0918	7.223	0.0100	6.797	1.215	0.3207	52	0.35	008.08	0
008	Fiber, Acid Detergent (%)	0001	7.370	0.0600	6.797	1.215	0.3207	52	0.47	008.08	0
008	Fiber, Acid Detergent (%)	0689	7.450	0.5000	6.797	1.215	0.3207	52	0.54	008.02	0
008	Fiber, Acid Detergent (%)	0726	7.705	1.110	6.797	1.215	0.3207	52	0.75	008.08	0
008	Fiber, Acid Detergent (%)	0226	7.735	0.1300	6.797	1.215	0.3207	52	0.77	008.02	0
008	Fiber, Acid Detergent (%)	0037	7.770	0.4000	6.797	1.215	0.3207	52	0.80	008.08	0
008	Fiber, Acid Detergent (%)	2193	7.810	0.1000	6.797	1.215	0.3207	52	0.83	008.08	0
008	Fiber, Acid Detergent (%)	0164	8.150	0.5000	6.797	1.215	0.3207	52	1.11	008.08	0
008	Fiber, Acid Detergent (%)	0510	8.150	0.5000	6.797	1.215	0.3207	52	1.11	008.08	0
008	Fiber, Acid Detergent (%)	0511	8.250	0.7000	6.797	1.215	0.3207	52	1.20	008.08	0
008	Fiber, Acid Detergent (%)	2012	8.315	0.4500	6.797	1.215	0.3207	52	1.25	008.08	0
008	Fiber, Acid Detergent (%)	0536	8.485	0.1100	6.797	1.215	0.3207	52	1.39	008.08	0
008	Fiber, Acid Detergent (%)	0278	8.550	0.3000	6.797	1.215	0.3207	52	1.44	008.08	0
008	Fiber, Acid Detergent (%)	0294	9.000	0.0000	6.797	1.215	0.3207	52	1.81	008.08	0
008	Fiber, Acid Detergent (%)	0693	10.44	0.9810	6.797	1.215	0.3207	52	3.00	008.08	0
008	Fiber, Acid Detergent (%)	0045	11.50	1.000	6.797	1.215	0.3207	52	3.87	008.08	0
008	Fiber, Acid Detergent (%)	0948	7.085	1.613	6.797	1.215	0.3207	52	0.24	008.08	1
009	Fiber, Neutral Detergent (%)	0353	10.60	0.1600	14.28	2.077	0.4453	51	-1.77	009.07	0
009	Fiber, Neutral Detergent (%)	0956	10.65	0.3000	14.28	2.077	0.4453	51	-1.75	009.09	0
009	Fiber, Neutral Detergent (%)	0878	10.89	0.1700	14.28	2.077	0.4453	51	-1.63	009.09	0
009	Fiber, Neutral Detergent (%)	0098	11.15	0.1000	14.28	2.077	0.4453	51	-1.51	009.07	0
009	Fiber, Neutral Detergent (%)	0263	11.88	0.0870	14.28	2.077	0.4453	51	-1.15	009.09	0
009	Fiber, Neutral Detergent (%)	0354	11.94	0.0400	14.28	2.077	0.4453	51	-1.12	009.09	0
009	Fiber, Neutral Detergent (%)	0910	12.40	0.0000	14.28	2.077	0.4453	51	-0.90	009.07	0
009	Fiber, Neutral Detergent (%)	0598	12.45	0.0500	14.28	2.077	0.4453	51	-0.88	009.09	0
009	Fiber, Neutral Detergent (%)	0357	12.50	0.4000	14.28	2.077	0.4453	51	-0.86	009.09	0
009	Fiber, Neutral Detergent (%)	0939	12.50	0.1800	14.28	2.077	0.4453	51	-0.86	009.09	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
009	Fiber, Neutral Detergent (%)	2136	12.50	0.8000	14.28	2.077	0.4453	51	-0.86	009.09	0
009	Fiber, Neutral Detergent (%)	0581	12.72	0.5300	14.28	2.077	0.4453	51	-0.75	009.09	0
009	Fiber, Neutral Detergent (%)	0948	12.80	0.2200	14.28	2.077	0.4453	51	-0.71	009.07	0
009	Fiber, Neutral Detergent (%)	0870	13.07	0.8165	14.28	2.077	0.4453	51	-0.58	009.09	0
009	Fiber, Neutral Detergent (%)	0358	13.13	1.690	14.28	2.077	0.4453	51	-0.55	009.09	0
009	Fiber, Neutral Detergent (%)	0226	13.33	0.6400	14.28	2.077	0.4453	51	-0.46	009.07	0
009	Fiber, Neutral Detergent (%)	0638	13.40	0.0000	14.28	2.077	0.4453	51	-0.42	009.09	0
009	Fiber, Neutral Detergent (%)	0345	13.42	0.2300	14.28	2.077	0.4453	51	-0.41	009.07	0
009	Fiber, Neutral Detergent (%)	0968	13.51	0.2350	14.28	2.077	0.4453	51	-0.37	009.09	0
009	Fiber, Neutral Detergent (%)	2192	13.56	0.0600	14.28	2.077	0.4453	51	-0.34	009.07	0
009	Fiber, Neutral Detergent (%)	0066	13.60	0.3100	14.28	2.077	0.4453	51	-0.33	009.09	0
009	Fiber, Neutral Detergent (%)	0294	13.75	0.3000	14.28	2.077	0.4453	51	-0.25	009.09	0
009	Fiber, Neutral Detergent (%)	0148	13.80	0.2000	14.28	2.077	0.4453	51	-0.23	009.09	0
009	Fiber, Neutral Detergent (%)	0083	13.93	0.3700	14.28	2.077	0.4453	51	-0.17	009.09	0
009	Fiber, Neutral Detergent (%)	0309	13.97	0.2788	14.28	2.077	0.4453	51	-0.15	009.07	0
009	Fiber, Neutral Detergent (%)	0693	14.02	0.9570	14.28	2.077	0.4453	51	-0.12	009.09	0
009	Fiber, Neutral Detergent (%)	0646	14.29	0.9600	14.28	2.077	0.4453	51	0.01	009.09	0
009	Fiber, Neutral Detergent (%)	0297	14.38	1.240	14.28	2.077	0.4453	51	0.05	009.09	0
009	Fiber, Neutral Detergent (%)	0610	14.55	0.5000	14.28	2.077	0.4453	51	0.13	009.07	0
009	Fiber, Neutral Detergent (%)	0889	14.66	0.0900	14.28	2.077	0.4453	51	0.18	009.99	0
009	Fiber, Neutral Detergent (%)	0278	14.70	1.400	14.28	2.077	0.4453	51	0.20	009.09	0
009	Fiber, Neutral Detergent (%)	2196	14.71	0.0000	14.28	2.077	0.4453	51	0.21	009.09	0
009	Fiber, Neutral Detergent (%)	2129	14.71	0.4350	14.28	2.077	0.4453	51	0.21	009.09	0
009	Fiber, Neutral Detergent (%)	0689	15.00	0.8000	14.28	2.077	0.4453	51	0.35	009.07	0
009	Fiber, Neutral Detergent (%)	0619	15.10	0.2000	14.28	2.077	0.4453	51	0.40	009.09	0
009	Fiber, Neutral Detergent (%)	0510	15.15	0.1000	14.28	2.077	0.4453	51	0.42	009.09	0
009	Fiber, Neutral Detergent (%)	0037	15.20	0.3700	14.28	2.077	0.4453	51	0.44	009.09	0
009	Fiber, Neutral Detergent (%)	0948	15.37	0.6320	14.28	2.077	0.4453	51	0.53	009.09	0
009	Fiber, Neutral Detergent (%)	2193	15.57	0.3800	14.28	2.077	0.4453	51	0.62	009.09	0
009	Fiber, Neutral Detergent (%)	2012	15.82	0.9200	14.28	2.077	0.4453	51	0.74	009.09	0
009	Fiber, Neutral Detergent (%)	0653	15.84	0.0100	14.28	2.077	0.4453	51	0.75	009.09	0
009	Fiber, Neutral Detergent (%)	0918	16.01	0.0100	14.28	2.077	0.4453	51	0.83	009.09	0
009	Fiber, Neutral Detergent (%)	0164	16.50	0.4000	14.28	2.077	0.4453	51	1.07	009.09	0
009	Fiber, Neutral Detergent (%)	0265	16.50	0.2000	14.28	2.077	0.4453	51	1.07	009.09	0
009	Fiber, Neutral Detergent (%)	0675	16.88	0.1400	14.28	2.077	0.4453	51	1.25	009.09	0
009	Fiber, Neutral Detergent (%)	0190	17.74	0.2800	14.28	2.077	0.4453	51	1.67	009.09	0
009	Fiber, Neutral Detergent (%)	0590	17.85	0.7000	14.28	2.077	0.4453	51	1.72	009.09	0
009	Fiber, Neutral Detergent (%)	0848	17.96	0.9800	14.28	2.077	0.4453	51	1.77	009.09	0
009	Fiber, Neutral Detergent (%)	0536	19.36	0.6500	14.28	2.077	0.4453	51	2.45	009.09	0
009	Fiber, Neutral Detergent (%)	0726	20.41	0.8900	14.28	2.077	0.4453	51	2.95	009.09	0
009	Fiber, Neutral Detergent (%)	0045	20.55	1.300	14.28	2.077	0.4453	51	3.02	009.09	0
009	Fiber, Neutral Detergent (%)	0407	12.71	3.264	14.28	2.077	0.4453	51	-0.75	009.09	1

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
010	Moisture (%)	0620	6.307	0.0328	7.600	0.3882	0.0948	22	-3.33	010.99	0
010	Moisture (%)	0553	6.475	0.0300	7.600	0.3882	0.0948	22	-2.90	010.11	0
010	Moisture (%)	0964	7.169	0.0380	7.600	0.3882	0.0948	22	-1.11	010.99	0
010	Moisture (%)	0941	7.310	0.2600	7.600	0.3882	0.0948	22	-0.75	010.99	0
010	Moisture (%)	2129	7.352	0.0350	7.600	0.3882	0.0948	22	-0.64	010.99	0
010	Moisture (%)	0726	7.365	0.0300	7.600	0.3882	0.0948	22	-0.60	010.99	0
010	Moisture (%)	0948	7.395	0.0100	7.600	0.3882	0.0948	22	-0.53	010.99	0
010	Moisture (%)	0692	7.400	0.0000	7.600	0.3882	0.0948	22	-0.51	010.99	0
010	Moisture (%)	0843	7.425	0.3900	7.600	0.3882	0.0948	22	-0.45	010.03	0
010	Moisture (%)	0939	7.510	0.1200	7.600	0.3882	0.0948	22	-0.23	010.99	0
010	Moisture (%)	0884	7.550	0.1000	7.600	0.3882	0.0948	22	-0.13	010.99	0
010	Moisture (%)	0164	7.615	0.0900	7.600	0.3882	0.0948	22	0.04	010.03	0
010	Moisture (%)	2169	7.655	0.1100	7.600	0.3882	0.0948	22	0.14	010.99	0
010	Moisture (%)	2043	7.670	0.2000	7.600	0.3882	0.0948	22	0.18	010.99	0
010	Moisture (%)	0652	7.800	0.0000	7.600	0.3882	0.0948	22	0.52	010.99	0
010	Moisture (%)	2089	7.825	0.2500	7.600	0.3882	0.0948	22	0.58	010.99	0
010	Moisture (%)	2150	7.825	0.0700	7.600	0.3882	0.0948	22	0.58	010.99	0
010	Moisture (%)	2076	7.845	0.0700	7.600	0.3882	0.0948	22	0.63	010.99	0
010	Moisture (%)	0914	8.030	0.0400	7.600	0.3882	0.0948	22	1.11	010.99	0
010	Moisture (%)	0190	8.075	0.0100	7.600	0.3882	0.0948	22	1.22	010.99	0
010	Moisture (%)	0889	8.160	0.1800	7.600	0.3882	0.0948	22	1.44	010.11	0
010	Moisture (%)	0405	9.060	0.0200	7.600	0.3882	0.0948	22	3.76	010.99	0
011	Loss on Drying (%)	2232	6.465	0.1500	8.086	0.2573	0.0871	70	-6.30	011.01	0
011	Loss on Drying (%)	0520	6.985	0.0100	8.086	0.2573	0.0871	70	-4.28	011.01	0
011	Loss on Drying (%)	2207	7.300	0.0000	8.086	0.2573	0.0871	70	-3.05	011.01	0
011	Loss on Drying (%)	0685	7.375	0.0300	8.086	0.2573	0.0871	70	-2.76	011.01	0
011	Loss on Drying (%)	2304	7.450	0.1600	8.086	0.2573	0.0871	70	-2.47	011.99	0
011	Loss on Drying (%)	2083	7.480	0.1800	8.086	0.2573	0.0871	70	-2.36	011.02	0
011	Loss on Drying (%)	0682	7.620	0.0000	8.086	0.2573	0.0871	70	-1.81	011.01	0
011	Loss on Drying (%)	2303	7.717	0.1700	8.086	0.2573	0.0871	70	-1.43	011.99	0
011	Loss on Drying (%)	0554	7.745	0.1500	8.086	0.2573	0.0871	70	-1.33	011.01	0
011	Loss on Drying (%)	0643	7.750	0.3000	8.086	0.2573	0.0871	70	-1.31	011.01	0
011	Loss on Drying (%)	0298	7.760	0.0200	8.086	0.2573	0.0871	70	-1.27	011.01	0
011	Loss on Drying (%)	0857	7.800	0.0000	8.086	0.2573	0.0871	70	-1.11	011.99	0
011	Loss on Drying (%)	0968	7.819	0.0330	8.086	0.2573	0.0871	70	-1.04	011.01	0
011	Loss on Drying (%)	0425	7.875	0.0500	8.086	0.2573	0.0871	70	-0.82	011.01	0
011	Loss on Drying (%)	2162	7.875	0.0300	8.086	0.2573	0.0871	70	-0.82	011.01	0
011	Loss on Drying (%)	0175	7.900	0.2000	8.086	0.2573	0.0871	70	-0.72	011.01	0
011	Loss on Drying (%)	0539	7.910	0.1800	8.086	0.2573	0.0871	70	-0.68	011.01	0
011	Loss on Drying (%)	2168	7.920	0.1800	8.086	0.2573	0.0871	70	-0.64	011.01	0
011	Loss on Drying (%)	0942	7.925	0.4100	8.086	0.2573	0.0871	70	-0.63	011.02	0
011	Loss on Drying (%)	2033	7.935	0.0100	8.086	0.2573	0.0871	70	-0.59	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
011	Loss on Drying (%)	2307	7.955	0.0300	8.086	0.2573	0.0871	70	-0.51	011.01	0
011	Loss on Drying (%)	0953	7.975	0.0300	8.086	0.2573	0.0871	70	-0.43	011.01	0
011	Loss on Drying (%)	2016	7.980	0.3400	8.086	0.2573	0.0871	70	-0.41	011.01	0
011	Loss on Drying (%)	0553	7.990	0.0200	8.086	0.2573	0.0871	70	-0.37	011.01	0
011	Loss on Drying (%)	0511	8.010	0.0400	8.086	0.2573	0.0871	70	-0.29	011.01	0
011	Loss on Drying (%)	0598	8.030	0.0200	8.086	0.2573	0.0871	70	-0.22	011.01	0
011	Loss on Drying (%)	0294	8.050	0.1000	8.086	0.2573	0.0871	70	-0.14	011.01	0
011	Loss on Drying (%)	0100	8.055	0.1500	8.086	0.2573	0.0871	70	-0.12	011.01	0
011	Loss on Drying (%)	2312	8.055	0.0300	8.086	0.2573	0.0871	70	-0.12	011.01	0
011	Loss on Drying (%)	0098	8.065	0.1300	8.086	0.2573	0.0871	70	-0.08	011.01	0
011	Loss on Drying (%)	0226	8.070	0.0600	8.086	0.2573	0.0871	70	-0.06	011.01	0
011	Loss on Drying (%)	0144	8.080	0.0200	8.086	0.2573	0.0871	70	-0.02	011.01	0
011	Loss on Drying (%)	0860	8.090	0.0000	8.086	0.2573	0.0871	70	0.02	011.01	0
011	Loss on Drying (%)	0574	8.100	0.0400	8.086	0.2573	0.0871	70	0.05	011.01	0
011	Loss on Drying (%)	2314	8.100	0.0000	8.086	0.2573	0.0871	70	0.05	011.01	0
011	Loss on Drying (%)	0843	8.110	0.1600	8.086	0.2573	0.0871	70	0.09	011.01	0
011	Loss on Drying (%)	0653	8.115	0.0700	8.086	0.2573	0.0871	70	0.11	011.01	0
011	Loss on Drying (%)	2174	8.130	0.0400	8.086	0.2573	0.0871	70	0.17	011.01	0
011	Loss on Drying (%)	0066	8.135	0.1100	8.086	0.2573	0.0871	70	0.19	011.01	0
011	Loss on Drying (%)	0309	8.145	0.0900	8.086	0.2573	0.0871	70	0.23	011.01	0
011	Loss on Drying (%)	2193	8.150	0.0400	8.086	0.2573	0.0871	70	0.25	011.01	0
011	Loss on Drying (%)	0859	8.158	0.0400	8.086	0.2573	0.0871	70	0.28	011.01	0
011	Loss on Drying (%)	0880	8.165	0.1900	8.086	0.2573	0.0871	70	0.31	011.01	0
011	Loss on Drying (%)	0164	8.170	0.1400	8.086	0.2573	0.0871	70	0.33	011.01	0
011	Loss on Drying (%)	2022	8.200	0.0000	8.086	0.2573	0.0871	70	0.44	011.01	0
011	Loss on Drying (%)	0354	8.205	0.0100	8.086	0.2573	0.0871	70	0.46	011.01	0
011	Loss on Drying (%)	0062	8.208	0.0250	8.086	0.2573	0.0871	70	0.47	011.01	0
011	Loss on Drying (%)	0242	8.245	0.1300	8.086	0.2573	0.0871	70	0.62	011.01	0
011	Loss on Drying (%)	0529	8.245	0.0100	8.086	0.2573	0.0871	70	0.62	011.02	0
011	Loss on Drying (%)	2188	8.250	0.0200	8.086	0.2573	0.0871	70	0.64	011.01	0
011	Loss on Drying (%)	2196	8.260	0.0000	8.086	0.2573	0.0871	70	0.68	011.01	0
011	Loss on Drying (%)	0263	8.279	0.1180	8.086	0.2573	0.0871	70	0.75	011.01	0
011	Loss on Drying (%)	0674	8.280	0.0000	8.086	0.2573	0.0871	70	0.75	011.01	0
011	Loss on Drying (%)	0618	8.285	0.0100	8.086	0.2573	0.0871	70	0.77	011.01	0
011	Loss on Drying (%)	0051	8.295	0.2500	8.086	0.2573	0.0871	70	0.81	011.01	0
011	Loss on Drying (%)	0265	8.300	0.0000	8.086	0.2573	0.0871	70	0.83	011.01	0
011	Loss on Drying (%)	0870	8.305	0.1129	8.086	0.2573	0.0871	70	0.85	011.01	0
011	Loss on Drying (%)	2073	8.310	0.0200	8.086	0.2573	0.0871	70	0.87	011.01	0
011	Loss on Drying (%)	0417	8.335	0.2500	8.086	0.2573	0.0871	70	0.97	011.02	0
011	Loss on Drying (%)	0536	8.345	0.0500	8.086	0.2573	0.0871	70	1.01	011.01	0
011	Loss on Drying (%)	0589	8.345	0.0900	8.086	0.2573	0.0871	70	1.01	011.01	0
011	Loss on Drying (%)	2006	8.350	0.0800	8.086	0.2573	0.0871	70	1.03	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
011	Loss on Drying (%)	0407	8.358	0.0862	8.086	0.2573	0.0871	70	1.06	011.01	0
011	Loss on Drying (%)	2268	8.360	0.0400	8.086	0.2573	0.0871	70	1.07	011.01	0
011	Loss on Drying (%)	0358	8.370	0.0400	8.086	0.2573	0.0871	70	1.10	011.01	0
011	Loss on Drying (%)	0123	8.400	0.0000	8.086	0.2573	0.0871	70	1.22	011.01	0
011	Loss on Drying (%)	0510	8.400	0.2000	8.086	0.2573	0.0871	70	1.22	011.01	0
011	Loss on Drying (%)	0202	8.425	0.3100	8.086	0.2573	0.0871	70	1.32	011.01	0
011	Loss on Drying (%)	2291	8.555	0.0700	8.086	0.2573	0.0871	70	1.82	011.01	0
011	Loss on Drying (%)	2181	8.630	0.0200	8.086	0.2573	0.0871	70	2.12	011.01	0
011	Loss on Drying (%)	0620	7.218	0.6466	8.086	0.2573	0.0871	70	-3.37	011.01	1
011	Loss on Drying (%)	0541	8.365	0.5700	8.086	0.2573	0.0871	70	1.08	011.01	1
012	Starch (%)	0297	14.61	0.0200	27.16	1.609	0.4808	36	-7.80	012.11	0
012	Starch (%)	0918	20.85	0.5640	27.16	1.609	0.4808	36	-3.92	012.01	0
012	Starch (%)	0510	23.35	0.1000	27.16	1.609	0.4808	36	-2.37	012.04	0
012	Starch (%)	2124	24.75	0.7000	27.16	1.609	0.4808	36	-1.50	012.03	0
012	Starch (%)	0956	25.05	0.9000	27.16	1.609	0.4808	36	-1.31	012.01	0
012	Starch (%)	0689	25.80	0.8000	27.16	1.609	0.4808	36	-0.84	012.00	0
012	Starch (%)	0870	26.00	0.8269	27.16	1.609	0.4808	36	-0.72	012.01	0
012	Starch (%)	2242	26.14	1.830	27.16	1.609	0.4808	36	-0.63	012.01	0
012	Starch (%)	0164	26.20	0.2000	27.16	1.609	0.4808	36	-0.59	012.00	0
012	Starch (%)	0164	26.25	0.7000	27.16	1.609	0.4808	36	-0.56	012.01	0
012	Starch (%)	2129	26.36	0.4910	27.16	1.609	0.4808	36	-0.49	012.04	0
012	Starch (%)	0638	26.40	0.0000	27.16	1.609	0.4808	36	-0.47	012.00	0
012	Starch (%)	0910	26.60	0.2000	27.16	1.609	0.4808	36	-0.35	012.01	0
012	Starch (%)	0676	26.65	0.3900	27.16	1.609	0.4808	36	-0.32	012.01	0
012	Starch (%)	0148	26.70	0.0000	27.16	1.609	0.4808	36	-0.28	012.01	0
012	Starch (%)	0045	27.05	0.5000	27.16	1.609	0.4808	36	-0.07	012.01	0
012	Starch (%)	0098	27.10	2.000	27.16	1.609	0.4808	36	-0.04	012.03	0
012	Starch (%)	0353	27.14	0.0500	27.16	1.609	0.4808	36	-0.01	012.20	0
012	Starch (%)	0693	27.24	0.3200	27.16	1.609	0.4808	36	0.05	012.01	0
012	Starch (%)	0278	27.30	0.8000	27.16	1.609	0.4808	36	0.09	012.04	0
012	Starch (%)	0407	27.31	0.3965	27.16	1.609	0.4808	36	0.09	012.03	0
012	Starch (%)	0683	27.51	0.1100	27.16	1.609	0.4808	36	0.22	012.00	0
012	Starch (%)	2188	27.55	0.0000	27.16	1.609	0.4808	36	0.24	012.00	0
012	Starch (%)	0619	27.80	0.2000	27.16	1.609	0.4808	36	0.40	012.00	0
012	Starch (%)	0948	27.95	0.2100	27.16	1.609	0.4808	36	0.49	012.04	0
012	Starch (%)	2006	27.98	0.1900	27.16	1.609	0.4808	36	0.51	012.00	0
012	Starch (%)	0354	28.01	0.3400	27.16	1.609	0.4808	36	0.53	012.00	0
012	Starch (%)	2196	28.20	0.0000	27.16	1.609	0.4808	36	0.65	012.00	0
012	Starch (%)	0878	28.28	0.5300	27.16	1.609	0.4808	36	0.70	012.00	0
012	Starch (%)	2193	28.50	0.1800	27.16	1.609	0.4808	36	0.84	012.00	0
012	Starch (%)	0910	28.60	0.6000	27.16	1.609	0.4808	36	0.90	012.00	0
012	Starch (%)	2192	28.74	0.0800	27.16	1.609	0.4808	36	0.98	012.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
012	Starch (%)	0297	30.51	2.000	27.16	1.609	0.4808	36	2.08	012.03	0
012	Starch (%)	0265	34.70	0.6000	27.16	1.609	0.4808	36	4.69	012.01	0
012	Starch (%)	0889	35.09	0.0600	27.16	1.609	0.4808	36	4.93	012.03	0
012	Starch (%)	0889	40.08	0.4200	27.16	1.609	0.4808	36	8.03	012.11	0
012	Starch (%)	2089	26.09	9.830	27.16	1.609	0.4808	36	-0.67	012.01	1
013	Fat, Acid Pretreat (%)	0853	7.750	0.1000	10.79	0.5113	0.1675	78	-5.95	013.02	0
013	Fat, Acid Pretreat (%)	0263	8.645	0.2140	10.79	0.5113	0.1675	78	-4.20	013.10	0
013	Fat, Acid Pretreat (%)	0618	8.679	0.0960	10.79	0.5113	0.1675	78	-4.14	013.08	0
013	Fat, Acid Pretreat (%)	0297	9.465	0.2700	10.79	0.5113	0.1675	78	-2.60	013.00	0
013	Fat, Acid Pretreat (%)	0939	9.830	0.6200	10.79	0.5113	0.1675	78	-1.88	013.10	0
013	Fat, Acid Pretreat (%)	2162	9.930	0.3200	10.79	0.5113	0.1675	78	-1.69	013.10	0
013	Fat, Acid Pretreat (%)	0845	9.990	0.0200	10.79	0.5113	0.1675	78	-1.57	013.10	0
013	Fat, Acid Pretreat (%)	2073	10.06	0.3500	10.79	0.5113	0.1675	78	-1.44	013.00	0
013	Fat, Acid Pretreat (%)	0554	10.11	0.2800	10.79	0.5113	0.1675	78	-1.34	013.13	0
013	Fat, Acid Pretreat (%)	0638	10.15	0.1000	10.79	0.5113	0.1675	78	-1.26	013.10	0
013	Fat, Acid Pretreat (%)	2146	10.18	0.0700	10.79	0.5113	0.1675	78	-1.21	013.00	0
013	Fat, Acid Pretreat (%)	0353	10.24	0.2600	10.79	0.5113	0.1675	78	-1.08	013.10	0
013	Fat, Acid Pretreat (%)	2187	10.30	0.0300	10.79	0.5113	0.1675	78	-0.98	013.00	0
013	Fat, Acid Pretreat (%)	0242	10.33	0.3000	10.79	0.5113	0.1675	78	-0.91	013.00	0
013	Fat, Acid Pretreat (%)	0036	10.35	0.1900	10.79	0.5113	0.1675	78	-0.87	013.10	0
013	Fat, Acid Pretreat (%)	0541	10.35	0.6400	10.79	0.5113	0.1675	78	-0.87	013.00	0
013	Fat, Acid Pretreat (%)	2314	10.40	0.0000	10.79	0.5113	0.1675	78	-0.77	013.02	0
013	Fat, Acid Pretreat (%)	0878	10.41	0.1400	10.79	0.5113	0.1675	78	-0.75	013.13	0
013	Fat, Acid Pretreat (%)	2124	10.42	0.0900	10.79	0.5113	0.1675	78	-0.74	013.10	0
013	Fat, Acid Pretreat (%)	0015	10.42	0.0500	10.79	0.5113	0.1675	78	-0.74	013.13	0
013	Fat, Acid Pretreat (%)	2076	10.48	0.0298	10.79	0.5113	0.1675	78	-0.61	013.00	0
013	Fat, Acid Pretreat (%)	2219	10.50	0.1500	10.79	0.5113	0.1675	78	-0.58	013.00	0
013	Fat, Acid Pretreat (%)	2196	10.54	0.0000	10.79	0.5113	0.1675	78	-0.50	013.10	0
013	Fat, Acid Pretreat (%)	0265	10.55	0.1000	10.79	0.5113	0.1675	78	-0.48	013.13	0
013	Fat, Acid Pretreat (%)	0019	10.56	0.0000	10.79	0.5113	0.1675	78	-0.46	013.13	0
013	Fat, Acid Pretreat (%)	0581	10.59	0.0300	10.79	0.5113	0.1675	78	-0.41	013.13	0
013	Fat, Acid Pretreat (%)	0964	10.60	0.3000	10.79	0.5113	0.1675	78	-0.38	013.02	0
013	Fat, Acid Pretreat (%)	2293	10.60	0.2800	10.79	0.5113	0.1675	78	-0.38	013.00	0
013	Fat, Acid Pretreat (%)	0653	10.71	0.0600	10.79	0.5113	0.1675	78	-0.16	013.13	0
013	Fat, Acid Pretreat (%)	0870	10.71	0.0151	10.79	0.5113	0.1675	78	-0.16	013.02	0
013	Fat, Acid Pretreat (%)	0034	10.72	0.3400	10.79	0.5113	0.1675	78	-0.14	013.00	0
013	Fat, Acid Pretreat (%)	0910	10.73	0.0500	10.79	0.5113	0.1675	78	-0.13	013.00	0
013	Fat, Acid Pretreat (%)	0003	10.74	0.7000	10.79	0.5113	0.1675	78	-0.10	013.13	0
013	Fat, Acid Pretreat (%)	0689	10.75	0.1000	10.79	0.5113	0.1675	78	-0.09	013.00	0
013	Fat, Acid Pretreat (%)	0610	10.75	0.1000	10.79	0.5113	0.1675	78	-0.09	013.10	0
013	Fat, Acid Pretreat (%)	0098	10.76	0.4700	10.79	0.5113	0.1675	78	-0.08	013.02	0
013	Fat, Acid Pretreat (%)	2312	10.76	0.3400	10.79	0.5113	0.1675	78	-0.07	013.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
013	Fat, Acid Pretreat (%)	0968	10.76	0.0880	10.79	0.5113	0.1675	78	-0.07	013.13	0
013	Fat, Acid Pretreat (%)	2016	10.77	0.5900	10.79	0.5113	0.1675	78	-0.06	013.00	0
013	Fat, Acid Pretreat (%)	2252	10.77	0.1400	10.79	0.5113	0.1675	78	-0.05	013.00	0
013	Fat, Acid Pretreat (%)	0843	10.83	0.2900	10.79	0.5113	0.1675	78	0.06	013.13	0
013	Fat, Acid Pretreat (%)	0309	10.83	0.0842	10.79	0.5113	0.1675	78	0.08	013.00	0
013	Fat, Acid Pretreat (%)	0407	10.84	0.0382	10.79	0.5113	0.1675	78	0.09	013.13	0
013	Fat, Acid Pretreat (%)	0652	10.85	0.1000	10.79	0.5113	0.1675	78	0.11	013.00	0
013	Fat, Acid Pretreat (%)	0948	10.88	0.1000	10.79	0.5113	0.1675	78	0.17	013.00	0
013	Fat, Acid Pretreat (%)	2156	10.93	0.0200	10.79	0.5113	0.1675	78	0.27	013.00	0
013	Fat, Acid Pretreat (%)	2220	10.94	0.0100	10.79	0.5113	0.1675	78	0.28	013.00	0
013	Fat, Acid Pretreat (%)	0675	10.94	0.4200	10.79	0.5113	0.1675	78	0.29	013.02	0
013	Fat, Acid Pretreat (%)	0884	10.95	0.1000	10.79	0.5113	0.1675	78	0.31	013.00	0
013	Fat, Acid Pretreat (%)	0100	10.96	0.0900	10.79	0.5113	0.1675	78	0.32	013.02	0
013	Fat, Acid Pretreat (%)	0017	11.00	0.2000	10.79	0.5113	0.1675	78	0.40	013.02	0
013	Fat, Acid Pretreat (%)	0682	11.02	0.0000	10.79	0.5113	0.1675	78	0.44	013.02	0
013	Fat, Acid Pretreat (%)	0065	11.05	0.2600	10.79	0.5113	0.1675	78	0.50	013.02	0
013	Fat, Acid Pretreat (%)	0510	11.05	0.3000	10.79	0.5113	0.1675	78	0.50	013.02	0
013	Fat, Acid Pretreat (%)	2053	11.10	0.0000	10.79	0.5113	0.1675	78	0.60	013.00	0
013	Fat, Acid Pretreat (%)	0139	11.16	0.1800	10.79	0.5113	0.1675	78	0.72	013.00	0
013	Fat, Acid Pretreat (%)	0148	11.19	0.0300	10.79	0.5113	0.1675	78	0.77	013.02	0
013	Fat, Acid Pretreat (%)	2264	11.19	0.0813	10.79	0.5113	0.1675	78	0.78	013.02	0
013	Fat, Acid Pretreat (%)	0354	11.20	0.0200	10.79	0.5113	0.1675	78	0.79	013.02	0
013	Fat, Acid Pretreat (%)	0643	11.20	0.1400	10.79	0.5113	0.1675	78	0.79	013.02	0
013	Fat, Acid Pretreat (%)	0876	11.20	0.4000	10.79	0.5113	0.1675	78	0.79	013.13	0
013	Fat, Acid Pretreat (%)	2307	11.21	0.0100	10.79	0.5113	0.1675	78	0.80	013.13	0
013	Fat, Acid Pretreat (%)	0035	11.21	0.0600	10.79	0.5113	0.1675	78	0.81	013.13	0
013	Fat, Acid Pretreat (%)	2150	11.22	0.1200	10.79	0.5113	0.1675	78	0.83	013.00	0
013	Fat, Acid Pretreat (%)	0051	11.22	0.1800	10.79	0.5113	0.1675	78	0.83	013.02	0
013	Fat, Acid Pretreat (%)	0035	11.26	0.0500	10.79	0.5113	0.1675	78	0.90	013.00	0
013	Fat, Acid Pretreat (%)	0164	11.27	0.0400	10.79	0.5113	0.1675	78	0.93	013.02	0
013	Fat, Acid Pretreat (%)	0202	11.32	0.2100	10.79	0.5113	0.1675	78	1.02	013.00	0
013	Fat, Acid Pretreat (%)	0553	11.34	0.0200	10.79	0.5113	0.1675	78	1.07	013.02	0
013	Fat, Acid Pretreat (%)	2259	11.35	0.0450	10.79	0.5113	0.1675	78	1.09	013.02	0
013	Fat, Acid Pretreat (%)	2304	11.35	0.0400	10.79	0.5113	0.1675	78	1.09	013.99	0
013	Fat, Acid Pretreat (%)	0914	11.45	0.0400	10.79	0.5113	0.1675	78	1.28	013.02	0
013	Fat, Acid Pretreat (%)	0407	11.47	0.0140	10.79	0.5113	0.1675	78	1.32	013.02	0
013	Fat, Acid Pretreat (%)	0676	11.50	0.3900	10.79	0.5113	0.1675	78	1.37	013.02	0
013	Fat, Acid Pretreat (%)	0921	11.64	0.2100	10.79	0.5113	0.1675	78	1.65	013.02	0
013	Fat, Acid Pretreat (%)	0425	11.97	0.0300	10.79	0.5113	0.1675	78	2.29	013.02	0
013	Fat, Acid Pretreat (%)	0083	13.03	0.3000	10.79	0.5113	0.1675	78	4.37	013.00	0
013	Fat, Acid Pretreat (%)	0618	13.27	0.4169	10.79	0.5113	0.1675	78	4.84	013.00	0
013	Fat, Acid Pretreat (%)	0511	9.165	1.450	10.79	0.5113	0.1675	78	-3.19	013.13	1

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
015	Aluminum (ppm)	0037	13.55	0.9000	44.66	10.51	1.786	15	-2.96	015.42	0
015	Aluminum (ppm)	0164	28.00	0.0000	44.66	10.51	1.786	15	-1.59	015.41	0
015	Aluminum (ppm)	0297	32.00	4.000	44.66	10.51	1.786	15	-1.20	015.43	0
015	Aluminum (ppm)	0510	40.50	1.000	44.66	10.51	1.786	15	-0.40	015.43	0
015	Aluminum (ppm)	0407	41.15	0.3126	44.66	10.51	1.786	15	-0.33	015.41	0
015	Aluminum (ppm)	0425	41.48	0.4900	44.66	10.51	1.786	15	-0.30	015.43	0
015	Aluminum (ppm)	0169	44.20	3.200	44.66	10.51	1.786	15	-0.04	015.43	0
015	Aluminum (ppm)	0918	45.58	0.3430	44.66	10.51	1.786	15	0.09	015.43	0
015	Aluminum (ppm)	0407	46.39	1.151	44.66	10.51	1.786	15	0.16	015.43	0
015	Aluminum (ppm)	0345	48.90	1.000	44.66	10.51	1.786	15	0.40	015.43	0
015	Aluminum (ppm)	0015	52.07	0.4500	44.66	10.51	1.786	15	0.71	015.53	0
015	Aluminum (ppm)	0553	53.35	1.900	44.66	10.51	1.786	15	0.83	015.53	0
015	Aluminum (ppm)	0520	54.85	0.5200	44.66	10.51	1.786	15	0.97	015.41	0
015	Aluminum (ppm)	0353	55.50	1.000	44.66	10.51	1.786	15	1.03	015.43	0
015	Aluminum (ppm)	0964	56.11	10.53	44.66	10.51	1.786	15	1.09	015.43	0
015	Aluminum (ppm)	2129	49.13	12.52	44.66	10.51	1.786	15	0.43	015.42	1
017	Boron (ppm)	0294	8.100	0.0000	9.778	1.031	0.5357	16	-1.63	017.42	0
017	Boron (ppm)	0226	8.730	0.0400	9.778	1.031	0.5357	16	-1.02	017.41	0
017	Boron (ppm)	0345	9.025	0.1500	9.778	1.031	0.5357	16	-0.73	017.43	0
017	Boron (ppm)	0037	9.200	0.2000	9.778	1.031	0.5357	16	-0.56	017.42	0
017	Boron (ppm)	0407	9.323	0.2219	9.778	1.031	0.5357	16	-0.44	017.41	0
017	Boron (ppm)	0083	9.400	0.2000	9.778	1.031	0.5357	16	-0.37	017.43	0
017	Boron (ppm)	0693	9.443	0.1180	9.778	1.031	0.5357	16	-0.32	017.42	0
017	Boron (ppm)	0045	9.505	0.9300	9.778	1.031	0.5357	16	-0.26	017.42	0
017	Boron (ppm)	0407	9.680	0.0910	9.778	1.031	0.5357	16	-0.10	017.43	0
017	Boron (ppm)	0918	9.700	0.0400	9.778	1.031	0.5357	16	-0.08	017.43	0
017	Boron (ppm)	0510	10.00	0.0000	9.778	1.031	0.5357	16	0.22	017.43	0
017	Boron (ppm)	0358	10.09	1.110	9.778	1.031	0.5357	16	0.30	017.41	0
017	Boron (ppm)	2129	10.15	0.2200	9.778	1.031	0.5357	16	0.36	017.42	0
017	Boron (ppm)	0353	12.00	2.000	9.778	1.031	0.5357	16	2.16	017.43	0
017	Boron (ppm)	0265	12.25	2.700	9.778	1.031	0.5357	16	2.40	017.42	0
017	Boron (ppm)	0405	13.92	0.5500	9.778	1.031	0.5357	16	4.01	017.44	0
017	Boron (ppm)	0015	49.64	21.13	9.778	1.031	0.5357	16	38.66	017.53	2
017	Boron (ppm)	0297	0.0000	0.0000	9.778	1.031	0.5357	16		017.43	4
019	Calcium (%)	0536	0.9755	0.0510	1.237	0.0820	0.0350	126	-3.19	019.02	0
019	Calcium (%)	0042	1.050	0.1000	1.237	0.0820	0.0350	126	-2.28	019.42	0
019	Calcium (%)	0175	1.085	0.0300	1.237	0.0820	0.0350	126	-1.85	019.31	0
019	Calcium (%)	0144	1.090	0.0800	1.237	0.0820	0.0350	126	-1.79	019.41	0
019	Calcium (%)	2129	1.093	0.0200	1.237	0.0820	0.0350	126	-1.76	019.42	0
019	Calcium (%)	0242	1.095	0.0300	1.237	0.0820	0.0350	126	-1.73	019.99	0
019	Calcium (%)	2053	1.100	0.0000	1.237	0.0820	0.0350	126	-1.67	019.42	0
019	Calcium (%)	0590	1.105	0.0100	1.237	0.0820	0.0350	126	-1.61	019.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
019	Calcium (%)	0154	1.106	0.0044	1.237	0.0820	0.0350	126	-1.60	019.52	0
019	Calcium (%)	0910	1.110	0.0400	1.237	0.0820	0.0350	126	-1.55	019.41	0
019	Calcium (%)	0536	1.120	0.1200	1.237	0.0820	0.0350	126	-1.43	019.31	0
019	Calcium (%)	0123	1.120	0.0800	1.237	0.0820	0.0350	126	-1.43	019.41	0
019	Calcium (%)	0868	1.140	0.0400	1.237	0.0820	0.0350	126	-1.18	019.99	0
019	Calcium (%)	0674	1.150	0.0000	1.237	0.0820	0.0350	126	-1.06	019.31	0
019	Calcium (%)	0003	1.150	0.0600	1.237	0.0820	0.0350	126	-1.06	019.41	0
019	Calcium (%)	0860	1.154	0.0000	1.237	0.0820	0.0350	126	-1.01	019.43	0
019	Calcium (%)	2055	1.154	0.0480	1.237	0.0820	0.0350	126	-1.01	019.53	0
019	Calcium (%)	2146	1.155	0.0100	1.237	0.0820	0.0350	126	-1.00	019.31	0
019	Calcium (%)	0425	1.155	0.0100	1.237	0.0820	0.0350	126	-1.00	019.43	0
019	Calcium (%)	2259	1.161	0.0840	1.237	0.0820	0.0350	126	-0.93	019.41	0
019	Calcium (%)	0036	1.161	0.0704	1.237	0.0820	0.0350	126	-0.92	019.43	0
019	Calcium (%)	0169	1.165	0.0300	1.237	0.0820	0.0350	126	-0.88	019.32	0
019	Calcium (%)	0878	1.170	0.0000	1.237	0.0820	0.0350	126	-0.82	019.00	0
019	Calcium (%)	0939	1.170	0.0000	1.237	0.0820	0.0350	126	-0.82	019.53	0
019	Calcium (%)	0142	1.172	0.0275	1.237	0.0820	0.0350	126	-0.79	019.31	0
019	Calcium (%)	0051	1.173	0.0110	1.237	0.0820	0.0350	126	-0.78	019.41	0
019	Calcium (%)	2109	1.175	0.0190	1.237	0.0820	0.0350	126	-0.76	019.00	0
019	Calcium (%)	0297	1.175	0.0500	1.237	0.0820	0.0350	126	-0.76	019.43	0
019	Calcium (%)	0918	1.177	0.0230	1.237	0.0820	0.0350	126	-0.74	019.43	0
019	Calcium (%)	0018	1.180	0.0200	1.237	0.0820	0.0350	126	-0.69	019.41	0
019	Calcium (%)	0065	1.183	0.0050	1.237	0.0820	0.0350	126	-0.66	019.44	0
019	Calcium (%)	2022	1.185	0.0700	1.237	0.0820	0.0350	126	-0.63	019.00	0
019	Calcium (%)	2022	1.185	0.0300	1.237	0.0820	0.0350	126	-0.63	019.31	0
019	Calcium (%)	2187	1.185	0.0500	1.237	0.0820	0.0350	126	-0.63	019.43	0
019	Calcium (%)	2192	1.185	0.0100	1.237	0.0820	0.0350	126	-0.63	019.43	0
019	Calcium (%)	2114	1.185	0.0500	1.237	0.0820	0.0350	126	-0.63	019.44	0
019	Calcium (%)	2232	1.186	0.0920	1.237	0.0820	0.0350	126	-0.62	019.44	0
019	Calcium (%)	0590	1.190	0.0200	1.237	0.0820	0.0350	126	-0.57	019.08	0
019	Calcium (%)	0529	1.195	0.0100	1.237	0.0820	0.0350	126	-0.51	019.31	0
019	Calcium (%)	2219	1.195	0.0300	1.237	0.0820	0.0350	126	-0.51	019.44	0
019	Calcium (%)	0407	1.203	0.0073	1.237	0.0820	0.0350	126	-0.42	019.41	0
019	Calcium (%)	0263	1.203	0.0100	1.237	0.0820	0.0350	126	-0.41	019.42	0
019	Calcium (%)	0676	1.205	0.0100	1.237	0.0820	0.0350	126	-0.39	019.99	0
019	Calcium (%)	0038	1.205	0.0515	1.237	0.0820	0.0350	126	-0.39	019.31	0
019	Calcium (%)	0298	1.210	0.0200	1.237	0.0820	0.0350	126	-0.33	019.41	0
019	Calcium (%)	0619	1.210	0.1400	1.237	0.0820	0.0350	126	-0.33	019.41	0
019	Calcium (%)	0366	1.210	0.0400	1.237	0.0820	0.0350	126	-0.33	019.42	0
019	Calcium (%)	2293	1.215	0.0500	1.237	0.0820	0.0350	126	-0.27	019.44	0
019	Calcium (%)	0139	1.218	0.0050	1.237	0.0820	0.0350	126	-0.24	019.41	0
019	Calcium (%)	2252	1.220	0.0000	1.237	0.0820	0.0350	126	-0.21	019.44	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
019	Calcium (%)	0914	1.225	0.0300	1.237	0.0820	0.0350	126	-0.15	019.43	0
019	Calcium (%)	0148	1.225	0.0700	1.237	0.0820	0.0350	126	-0.14	019.41	0
019	Calcium (%)	2156	1.225	0.0100	1.237	0.0820	0.0350	126	-0.14	019.44	0
019	Calcium (%)	2146	1.225	0.0500	1.237	0.0820	0.0350	126	-0.14	019.51	0
019	Calcium (%)	2006	1.226	0.0100	1.237	0.0820	0.0350	126	-0.13	019.00	0
019	Calcium (%)	0693	1.228	0.0090	1.237	0.0820	0.0350	126	-0.11	019.42	0
019	Calcium (%)	0520	1.229	0.0690	1.237	0.0820	0.0350	126	-0.10	019.41	0
019	Calcium (%)	2009	1.229	0.0020	1.237	0.0820	0.0350	126	-0.10	019.08	0
019	Calcium (%)	0897	1.230	0.0000	1.237	0.0820	0.0350	126	-0.08	019.00	0
019	Calcium (%)	2196	1.230	0.0000	1.237	0.0820	0.0350	126	-0.08	019.08	0
019	Calcium (%)	0553	1.230	0.0400	1.237	0.0820	0.0350	126	-0.08	019.53	0
019	Calcium (%)	0968	1.232	0.0200	1.237	0.0820	0.0350	126	-0.06	019.43	0
019	Calcium (%)	0407	1.235	0.0070	1.237	0.0820	0.0350	126	-0.03	019.43	0
019	Calcium (%)	0868	1.235	0.0100	1.237	0.0820	0.0350	126	-0.02	019.31	0
019	Calcium (%)	0884	1.235	0.0300	1.237	0.0820	0.0350	126	-0.02	019.31	0
019	Calcium (%)	0019	1.235	0.0700	1.237	0.0820	0.0350	126	-0.02	019.41	0
019	Calcium (%)	0164	1.235	0.0100	1.237	0.0820	0.0350	126	-0.02	019.44	0
019	Calcium (%)	0653	1.238	0.0000	1.237	0.0820	0.0350	126	0.01	019.41	0
019	Calcium (%)	0939	1.240	0.0200	1.237	0.0820	0.0350	126	0.04	019.31	0
019	Calcium (%)	2196	1.240	0.0000	1.237	0.0820	0.0350	126	0.04	019.31	0
019	Calcium (%)	0010	1.240	0.0400	1.237	0.0820	0.0350	126	0.04	019.43	0
019	Calcium (%)	0036	1.240	0.0126	1.237	0.0820	0.0350	126	0.04	019.03	0
019	Calcium (%)	2006	1.241	0.0920	1.237	0.0820	0.0350	126	0.05	019.09	0
019	Calcium (%)	0878	1.245	0.0100	1.237	0.0820	0.0350	126	0.10	019.41	0
019	Calcium (%)	2012	1.245	0.0100	1.237	0.0820	0.0350	126	0.10	019.41	0
019	Calcium (%)	0035	1.245	0.0100	1.237	0.0820	0.0350	126	0.10	019.42	0
019	Calcium (%)	2220	1.245	0.0100	1.237	0.0820	0.0350	126	0.10	019.44	0
019	Calcium (%)	2193	1.250	0.0000	1.237	0.0820	0.0350	126	0.16	019.08	0
019	Calcium (%)	0687	1.250	0.1000	1.237	0.0820	0.0350	126	0.16	019.31	0
019	Calcium (%)	0265	1.250	0.2000	1.237	0.0820	0.0350	126	0.16	019.42	0
019	Calcium (%)	2314	1.250	0.1000	1.237	0.0820	0.0350	126	0.16	019.43	0
019	Calcium (%)	0358	1.255	0.0500	1.237	0.0820	0.0350	126	0.22	019.41	0
019	Calcium (%)	0083	1.255	0.0100	1.237	0.0820	0.0350	126	0.22	019.43	0
019	Calcium (%)	0017	1.259	0.0545	1.237	0.0820	0.0350	126	0.27	019.43	0
019	Calcium (%)	0948	1.265	0.0100	1.237	0.0820	0.0350	126	0.34	019.31	0
019	Calcium (%)	0345	1.268	0.0420	1.237	0.0820	0.0350	126	0.38	019.43	0
019	Calcium (%)	0870	1.270	0.0350	1.237	0.0820	0.0350	126	0.40	019.42	0
019	Calcium (%)	2295	1.275	0.0100	1.237	0.0820	0.0350	126	0.47	019.00	0
019	Calcium (%)	0066	1.277	0.0290	1.237	0.0820	0.0350	126	0.48	019.41	0
019	Calcium (%)	0848	1.280	0.0000	1.237	0.0820	0.0350	126	0.53	019.41	0
019	Calcium (%)	0682	1.280	0.0000	1.237	0.0820	0.0350	126	0.53	019.43	0
019	Calcium (%)	0098	1.285	0.0300	1.237	0.0820	0.0350	126	0.59	019.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
019	Calcium (%)	0610	1.285	0.0100	1.237	0.0820	0.0350	126	0.59	019.43	0
019	Calcium (%)	0045	1.295	0.0300	1.237	0.0820	0.0350	126	0.71	019.42	0
019	Calcium (%)	2124	1.295	0.0500	1.237	0.0820	0.0350	126	0.71	019.43	0
019	Calcium (%)	0683	1.300	0.0200	1.237	0.0820	0.0350	126	0.77	019.00	0
019	Calcium (%)	0689	1.300	0.0200	1.237	0.0820	0.0350	126	0.77	019.08	0
019	Calcium (%)	2089	1.300	0.0600	1.237	0.0820	0.0350	126	0.77	019.43	0
019	Calcium (%)	0015	1.300	0.0400	1.237	0.0820	0.0350	126	0.77	019.53	0
019	Calcium (%)	0357	1.301	0.0015	1.237	0.0820	0.0350	126	0.78	019.42	0
019	Calcium (%)	0278	1.305	0.0300	1.237	0.0820	0.0350	126	0.83	019.42	0
019	Calcium (%)	0032	1.309	0.0320	1.237	0.0820	0.0350	126	0.88	019.43	0
019	Calcium (%)	0226	1.310	0.0000	1.237	0.0820	0.0350	126	0.89	019.41	0
019	Calcium (%)	0354	1.310	0.0200	1.237	0.0820	0.0350	126	0.89	019.41	0
019	Calcium (%)	0100	1.310	0.0000	1.237	0.0820	0.0350	126	0.89	019.99	0
019	Calcium (%)	0009	1.319	0.0060	1.237	0.0820	0.0350	126	1.00	019.42	0
019	Calcium (%)	0405	1.322	0.0080	1.237	0.0820	0.0350	126	1.04	019.41	0
019	Calcium (%)	0035	1.330	0.0400	1.237	0.0820	0.0350	126	1.14	019.43	0
019	Calcium (%)	0572	1.330	0.1000	1.237	0.0820	0.0350	126	1.14	019.53	0
019	Calcium (%)	2115	1.335	0.0300	1.237	0.0820	0.0350	126	1.20	019.31	0
019	Calcium (%)	0202	1.335	0.0500	1.237	0.0820	0.0350	126	1.20	019.43	0
019	Calcium (%)	0692	1.340	0.0600	1.237	0.0820	0.0350	126	1.26	019.42	0
019	Calcium (%)	0037	1.356	0.0000	1.237	0.0820	0.0350	126	1.45	019.42	0
019	Calcium (%)	0001	1.361	0.0600	1.237	0.0820	0.0350	126	1.51	019.31	0
019	Calcium (%)	0598	1.371	0.0020	1.237	0.0820	0.0350	126	1.64	019.41	0
019	Calcium (%)	0294	1.375	0.0500	1.237	0.0820	0.0350	126	1.69	019.42	0
019	Calcium (%)	0098	1.390	0.0400	1.237	0.0820	0.0350	126	1.87	019.43	0
019	Calcium (%)	0353	1.410	0.0400	1.237	0.0820	0.0350	126	2.11	019.43	0
019	Calcium (%)	2207	1.420	0.2200	1.237	0.0820	0.0350	126	2.23	019.42	0
019	Calcium (%)	0620	1.428	0.1038	1.237	0.0820	0.0350	126	2.33	019.31	0
019	Calcium (%)	0726	1.430	0.0000	1.237	0.0820	0.0350	126	2.36	019.42	0
019	Calcium (%)	0685	1.435	0.0100	1.237	0.0820	0.0350	126	2.42	019.41	0
019	Calcium (%)	0190	1.440	0.0200	1.237	0.0820	0.0350	126	2.48	019.42	0
019	Calcium (%)	2188	1.470	0.0000	1.237	0.0820	0.0350	126	2.84	019.08	0
019	Calcium (%)	2234	1.490	0.1600	1.237	0.0820	0.0350	126	3.09	019.43	0
019	Calcium (%)	2181	1.560	0.0200	1.237	0.0820	0.0350	126	3.94	019.00	0
019	Calcium (%)	0510	126.0	4.000	1.237	0.0820	0.0350	126	1522.31	019.43	2
021	Cobalt (ppm)	0910	0.1500	0.0000	0.3263	0.1492	0.0613	14	-1.18	021.52	0
021	Cobalt (ppm)	0689	0.1500	0.1000	0.3263	0.1492	0.0613	14	-1.18	021.31	0
021	Cobalt (ppm)	0918	0.1612	0.0178	0.3263	0.1492	0.0613	14	-1.11	021.53	0
021	Cobalt (ppm)	2055	0.2620	0.0020	0.3263	0.1492	0.0613	14	-0.43	021.53	0
021	Cobalt (ppm)	0964	0.2725	0.0475	0.3263	0.1492	0.0613	14	-0.36	021.43	0
021	Cobalt (ppm)	0510	0.2850	0.0100	0.3263	0.1492	0.0613	14	-0.28	021.43	0
021	Cobalt (ppm)	0169	0.3000	0.0400	0.3263	0.1492	0.0613	14	-0.18	021.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
021	Cobalt (ppm)	0045	0.3100	0.2200	0.3263	0.1492	0.0613	14	-0.11	021.42	0
021	Cobalt (ppm)	0553	0.3385	0.0370	0.3263	0.1492	0.0613	14	0.08	021.53	0
021	Cobalt (ppm)	0939	0.3550	0.0900	0.3263	0.1492	0.0613	14	0.19	021.53	0
021	Cobalt (ppm)	0015	0.4200	0.1200	0.3263	0.1492	0.0613	14	0.63	021.53	0
021	Cobalt (ppm)	0407	0.4893	0.0182	0.3263	0.1492	0.0613	14	1.09	021.41	0
021	Cobalt (ppm)	0939	0.5250	0.0100	0.3263	0.1492	0.0613	14	1.33	021.31	0
021	Cobalt (ppm)	0407	0.7090	0.1460	0.3263	0.1492	0.0613	14	2.56	021.43	0
021	Cobalt (ppm)	0353	1.408	1.658	0.3263	0.1492	0.0613	14	7.25	021.43	1
021	Cobalt (ppm)	0693	0.0445	0.0890	0.3263	0.1492	0.0613	14		021.42	4
021	Cobalt (ppm)	0164	< 0.75		0.3263	0.1492	0.0613	14		021.31	5
021	Cobalt (ppm)	0682	< 1		0.3263	0.1492	0.0613	14		021.43	5
022	Copper (ppm)	2188	8.755	0.0000	14.02	1.437	0.7375	91	-3.66	022.31	0
022	Copper (ppm)	0536	9.025	0.3500	14.02	1.437	0.7375	91	-3.48	022.31	0
022	Copper (ppm)	0685	9.920	0.0200	14.02	1.437	0.7375	91	-2.85	022.42	0
022	Copper (ppm)	2314	10.25	1.100	14.02	1.437	0.7375	91	-2.62	022.43	0
022	Copper (ppm)	2055	11.50	1.000	14.02	1.437	0.7375	91	-1.75	022.53	0
022	Copper (ppm)	0015	11.70	3.580	14.02	1.437	0.7375	91	-1.62	022.53	0
022	Copper (ppm)	0653	11.82	0.0800	14.02	1.437	0.7375	91	-1.53	022.41	0
022	Copper (ppm)	0914	12.44	0.5080	14.02	1.437	0.7375	91	-1.10	022.53	0
022	Copper (ppm)	0032	12.45	0.4200	14.02	1.437	0.7375	91	-1.09	022.43	0
022	Copper (ppm)	0572	12.45	1.500	14.02	1.437	0.7375	91	-1.09	022.53	0
022	Copper (ppm)	0100	12.50	1.000	14.02	1.437	0.7375	91	-1.06	022.99	0
022	Copper (ppm)	0242	12.50	1.000	14.02	1.437	0.7375	91	-1.06	022.99	0
022	Copper (ppm)	0918	12.54	0.2000	14.02	1.437	0.7375	91	-1.03	022.43	0
022	Copper (ppm)	2129	12.55	0.8300	14.02	1.437	0.7375	91	-1.03	022.42	0
022	Copper (ppm)	2192	12.59	0.0000	14.02	1.437	0.7375	91	-1.00	022.43	0
022	Copper (ppm)	0045	12.70	0.0000	14.02	1.437	0.7375	91	-0.92	022.42	0
022	Copper (ppm)	0169	12.70	0.4000	14.02	1.437	0.7375	91	-0.92	022.43	0
022	Copper (ppm)	0407	12.92	0.0200	14.02	1.437	0.7375	91	-0.76	022.41	0
022	Copper (ppm)	0366	13.00	0.0000	14.02	1.437	0.7375	91	-0.71	022.42	0
022	Copper (ppm)	0083	13.00	0.0000	14.02	1.437	0.7375	91	-0.71	022.43	0
022	Copper (ppm)	0968	13.00	0.0000	14.02	1.437	0.7375	91	-0.71	022.43	0
022	Copper (ppm)	0345	13.05	0.1000	14.02	1.437	0.7375	91	-0.68	022.43	0
022	Copper (ppm)	0590	13.05	0.5000	14.02	1.437	0.7375	91	-0.68	022.99	0
022	Copper (ppm)	0904	13.06	0.7300	14.02	1.437	0.7375	91	-0.67	022.33	0
022	Copper (ppm)	0017	13.07	1.371	14.02	1.437	0.7375	91	-0.66	022.43	0
022	Copper (ppm)	0619	13.15	0.1000	14.02	1.437	0.7375	91	-0.61	022.41	0
022	Copper (ppm)	0939	13.15	0.4400	14.02	1.437	0.7375	91	-0.61	022.53	0
022	Copper (ppm)	0278	13.20	1.000	14.02	1.437	0.7375	91	-0.57	022.42	0
022	Copper (ppm)	0098	13.35	0.0900	14.02	1.437	0.7375	91	-0.47	022.43	0
022	Copper (ppm)	0263	13.38	0.1520	14.02	1.437	0.7375	91	-0.44	022.42	0
022	Copper (ppm)	0682	13.40	0.0000	14.02	1.437	0.7375	91	-0.43	022.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
022	Copper (ppm)	0553	13.40	0.2000	14.02	1.437	0.7375	91	-0.43	022.53	0
022	Copper (ppm)	0033	13.45	0.3000	14.02	1.437	0.7375	91	-0.40	022.43	0
022	Copper (ppm)	0407	13.49	0.0880	14.02	1.437	0.7375	91	-0.37	022.43	0
022	Copper (ppm)	0910	13.50	1.000	14.02	1.437	0.7375	91	-0.36	022.41	0
022	Copper (ppm)	0297	13.50	1.000	14.02	1.437	0.7375	91	-0.36	022.43	0
022	Copper (ppm)	0051	13.55	0.6300	14.02	1.437	0.7375	91	-0.33	022.41	0
022	Copper (ppm)	0265	13.65	0.1000	14.02	1.437	0.7375	91	-0.26	022.42	0
022	Copper (ppm)	0692	13.70	2.800	14.02	1.437	0.7375	91	-0.22	022.42	0
022	Copper (ppm)	0098	13.72	0.4800	14.02	1.437	0.7375	91	-0.21	022.41	0
022	Copper (ppm)	0037	13.80	0.2000	14.02	1.437	0.7375	91	-0.15	022.42	0
022	Copper (ppm)	0848	13.86	0.0300	14.02	1.437	0.7375	91	-0.12	022.41	0
022	Copper (ppm)	0689	13.90	1.000	14.02	1.437	0.7375	91	-0.08	022.31	0
022	Copper (ppm)	0038	13.95	0.3900	14.02	1.437	0.7375	91	-0.05	022.32	0
022	Copper (ppm)	2114	13.97	0.2200	14.02	1.437	0.7375	91	-0.04	022.44	0
022	Copper (ppm)	0870	13.97	0.1800	14.02	1.437	0.7375	91	-0.04	022.42	0
022	Copper (ppm)	0186	14.00	2.000	14.02	1.437	0.7375	91	-0.02	022.42	0
022	Copper (ppm)	0035	14.00	0.0000	14.02	1.437	0.7375	91	-0.02	022.43	0
022	Copper (ppm)	0186	14.00	2.000	14.02	1.437	0.7375	91	-0.02	022.52	0
022	Copper (ppm)	0939	14.04	0.9700	14.02	1.437	0.7375	91	0.01	022.31	0
022	Copper (ppm)	0202	14.04	0.3900	14.02	1.437	0.7375	91	0.01	022.43	0
022	Copper (ppm)	0148	14.04	0.1000	14.02	1.437	0.7375	91	0.01	022.41	0
022	Copper (ppm)	0693	14.23	1.001	14.02	1.437	0.7375	91	0.14	022.42	0
022	Copper (ppm)	0726	14.24	0.0100	14.02	1.437	0.7375	91	0.15	022.42	0
022	Copper (ppm)	2124	14.29	0.2800	14.02	1.437	0.7375	91	0.19	022.43	0
022	Copper (ppm)	0610	14.30	0.0000	14.02	1.437	0.7375	91	0.19	022.43	0
022	Copper (ppm)	0357	14.30	0.0620	14.02	1.437	0.7375	91	0.19	022.42	0
022	Copper (ppm)	0405	14.31	0.4400	14.02	1.437	0.7375	91	0.20	022.41	0
022	Copper (ppm)	0035	14.50	1.000	14.02	1.437	0.7375	91	0.33	022.42	0
022	Copper (ppm)	0510	14.50	1.000	14.02	1.437	0.7375	91	0.33	022.43	0
022	Copper (ppm)	0164	14.50	1.000	14.02	1.437	0.7375	91	0.33	022.44	0
022	Copper (ppm)	0009	14.51	0.9900	14.02	1.437	0.7375	91	0.34	022.42	0
022	Copper (ppm)	0884	14.65	0.1000	14.02	1.437	0.7375	91	0.44	022.31	0
022	Copper (ppm)	0294	14.66	0.2800	14.02	1.437	0.7375	91	0.44	022.42	0
022	Copper (ppm)	0948	14.80	0.3700	14.02	1.437	0.7375	91	0.54	022.33	0
022	Copper (ppm)	0190	14.89	0.1800	14.02	1.437	0.7375	91	0.60	022.42	0
022	Copper (ppm)	0010	14.90	0.6000	14.02	1.437	0.7375	91	0.61	022.33	0
022	Copper (ppm)	2146	14.99	1.810	14.02	1.437	0.7375	91	0.67	022.31	0
022	Copper (ppm)	0878	15.00	2.000	14.02	1.437	0.7375	91	0.68	022.41	0
022	Copper (ppm)	0598	15.01	0.1400	14.02	1.437	0.7375	91	0.69	022.41	0
022	Copper (ppm)	0042	15.05	3.700	14.02	1.437	0.7375	91	0.72	022.42	0
022	Copper (ppm)	2053	15.10	0.2000	14.02	1.437	0.7375	91	0.75	022.42	0
022	Copper (ppm)	0019	15.27	2.450	14.02	1.437	0.7375	91	0.86	022.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
022	Copper (ppm)	0358	15.30	2.330	14.02	1.437	0.7375	91	0.89	022.41	0
022	Copper (ppm)	0520	15.41	0.0100	14.02	1.437	0.7375	91	0.96	022.41	0
022	Copper (ppm)	2146	15.43	1.590	14.02	1.437	0.7375	91	0.98	022.51	0
022	Copper (ppm)	0964	15.51	1.703	14.02	1.437	0.7375	91	1.04	022.43	0
022	Copper (ppm)	0868	15.60	0.6100	14.02	1.437	0.7375	91	1.09	022.31	0
022	Copper (ppm)	0354	15.78	0.0400	14.02	1.437	0.7375	91	1.22	022.41	0
022	Copper (ppm)	2089	15.79	3.410	14.02	1.437	0.7375	91	1.23	022.43	0
022	Copper (ppm)	2207	16.00	2.000	14.02	1.437	0.7375	91	1.38	022.42	0
022	Copper (ppm)	0353	16.00	2.000	14.02	1.437	0.7375	91	1.38	022.43	0
022	Copper (ppm)	0226	16.03	0.2600	14.02	1.437	0.7375	91	1.40	022.41	0
022	Copper (ppm)	2196	16.85	0.0000	14.02	1.437	0.7375	91	1.97	022.31	0
022	Copper (ppm)	2115	16.88	0.4500	14.02	1.437	0.7375	91	1.98	022.31	0
022	Copper (ppm)	0529	16.95	1.900	14.02	1.437	0.7375	91	2.04	022.31	0
022	Copper (ppm)	2234	17.01	1.150	14.02	1.437	0.7375	91	2.08	022.43	0
022	Copper (ppm)	0175	18.00	0.0000	14.02	1.437	0.7375	91	2.77	022.31	0
022	Copper (ppm)	0003	18.00	0.0000	14.02	1.437	0.7375	91	2.77	022.41	0
022	Copper (ppm)	2012	18.67	0.4900	14.02	1.437	0.7375	91	3.23	022.41	0
022	Copper (ppm)	0066	19.69	0.9920	14.02	1.437	0.7375	91	3.94	022.44	0
022	Copper (ppm)	0620	153.5	82.98	14.02	1.437	0.7375	91	97.03	022.31	2
025	Iron (ppm)	0154	146.1	3.630	193.8	18.44	9.169	85	-2.59	025.52	0
025	Iron (ppm)	0620	150.8	24.89	193.8	18.44	9.169	85	-2.33	025.31	0
025	Iron (ppm)	0619	155.6	2.548	193.8	18.44	9.169	85	-2.07	025.41	0
025	Iron (ppm)	2114	156.9	3.000	193.8	18.44	9.169	85	-2.00	025.41	0
025	Iron (ppm)	0939	160.1	1.310	193.8	18.44	9.169	85	-1.83	025.31	0
025	Iron (ppm)	2314	163.1	3.600	193.8	18.44	9.169	85	-1.67	025.43	0
025	Iron (ppm)	0123	169.5	9.000	193.8	18.44	9.169	85	-1.32	025.41	0
025	Iron (ppm)	0653	173.6	5.200	193.8	18.44	9.169	85	-1.10	025.41	0
025	Iron (ppm)	2012	174.6	1.430	193.8	18.44	9.169	85	-1.05	025.41	0
025	Iron (ppm)	0226	175.0	7.230	193.8	18.44	9.169	85	-1.02	025.41	0
025	Iron (ppm)	2055	175.0	8.000	193.8	18.44	9.169	85	-1.02	025.53	0
025	Iron (ppm)	0051	175.5	3.000	193.8	18.44	9.169	85	-0.99	025.41	0
025	Iron (ppm)	2022	175.9	0.3500	193.8	18.44	9.169	85	-0.97	025.31	0
025	Iron (ppm)	0278	176.0	12.00	193.8	18.44	9.169	85	-0.97	025.42	0
025	Iron (ppm)	0529	176.2	3.500	193.8	18.44	9.169	85	-0.96	025.31	0
025	Iron (ppm)	0407	176.2	7.673	193.8	18.44	9.169	85	-0.96	025.41	0
025	Iron (ppm)	0868	177.5	3.000	193.8	18.44	9.169	85	-0.89	025.31	0
025	Iron (ppm)	2146	178.0	0.5200	193.8	18.44	9.169	85	-0.86	025.31	0
025	Iron (ppm)	0001	179.0	11.50	193.8	18.44	9.169	85	-0.81	025.31	0
025	Iron (ppm)	0242	179.0	12.00	193.8	18.44	9.169	85	-0.80	025.99	0
025	Iron (ppm)	0037	179.9	16.10	193.8	18.44	9.169	85	-0.76	025.42	0
025	Iron (ppm)	0354	180.1	1.220	193.8	18.44	9.169	85	-0.75	025.41	0
025	Iron (ppm)	0910	182.5	27.00	193.8	18.44	9.169	85	-0.61	025.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
025	Iron (ppm)	0169	182.5	1.000	193.8	18.44	9.169	85	-0.61	025.43	0
025	Iron (ppm)	0019	183.0	6.200	193.8	18.44	9.169	85	-0.59	025.41	0
025	Iron (ppm)	0042	183.5	5.000	193.8	18.44	9.169	85	-0.56	025.42	0
025	Iron (ppm)	0939	184.5	5.000	193.8	18.44	9.169	85	-0.51	025.53	0
025	Iron (ppm)	0878	185.0	4.000	193.8	18.44	9.169	85	-0.48	025.41	0
025	Iron (ppm)	0358	186.1	15.82	193.8	18.44	9.169	85	-0.42	025.41	0
025	Iron (ppm)	0345	186.5	3.000	193.8	18.44	9.169	85	-0.40	025.43	0
025	Iron (ppm)	2152	186.9	3.240	193.8	18.44	9.169	85	-0.37	025.41	0
025	Iron (ppm)	0265	187.5	21.00	193.8	18.44	9.169	85	-0.34	025.42	0
025	Iron (ppm)	0425	187.8	3.260	193.8	18.44	9.169	85	-0.33	025.43	0
025	Iron (ppm)	0148	188.5	1.000	193.8	18.44	9.169	85	-0.29	025.41	0
025	Iron (ppm)	0297	188.5	13.00	193.8	18.44	9.169	85	-0.29	025.43	0
025	Iron (ppm)	0407	188.9	1.534	193.8	18.44	9.169	85	-0.27	025.43	0
025	Iron (ppm)	0033	189.5	1.000	193.8	18.44	9.169	85	-0.24	025.43	0
025	Iron (ppm)	0968	189.5	3.000	193.8	18.44	9.169	85	-0.24	025.43	0
025	Iron (ppm)	0366	191.5	7.000	193.8	18.44	9.169	85	-0.13	025.42	0
025	Iron (ppm)	2129	191.7	2.000	193.8	18.44	9.169	85	-0.12	025.42	0
025	Iron (ppm)	0098	193.0	14.90	193.8	18.44	9.169	85	-0.05	025.41	0
025	Iron (ppm)	0610	193.5	9.000	193.8	18.44	9.169	85	-0.02	025.43	0
025	Iron (ppm)	0017	194.6	9.639	193.8	18.44	9.169	85	0.04	025.43	0
025	Iron (ppm)	0553	195.0	6.000	193.8	18.44	9.169	85	0.06	025.53	0
025	Iron (ppm)	0045	195.5	13.00	193.8	18.44	9.169	85	0.09	025.42	0
025	Iron (ppm)	0726	195.6	0.4000	193.8	18.44	9.169	85	0.10	025.42	0
025	Iron (ppm)	0918	195.7	5.860	193.8	18.44	9.169	85	0.10	025.43	0
025	Iron (ppm)	0263	195.8	2.983	193.8	18.44	9.169	85	0.10	025.42	0
025	Iron (ppm)	0693	196.3	6.009	193.8	18.44	9.169	85	0.13	025.42	0
025	Iron (ppm)	2196	196.9	0.0000	193.8	18.44	9.169	85	0.16	025.31	0
025	Iron (ppm)	0164	197.0	6.000	193.8	18.44	9.169	85	0.17	025.41	0
025	Iron (ppm)	0015	197.9	0.8800	193.8	18.44	9.169	85	0.22	025.53	0
025	Iron (ppm)	0083	198.0	10.00	193.8	18.44	9.169	85	0.23	025.43	0
025	Iron (ppm)	0100	198.0	4.000	193.8	18.44	9.169	85	0.23	025.99	0
025	Iron (ppm)	0510	198.5	1.000	193.8	18.44	9.169	85	0.25	025.43	0
025	Iron (ppm)	0294	199.5	13.00	193.8	18.44	9.169	85	0.31	025.42	0
025	Iron (ppm)	0692	200.0	32.00	193.8	18.44	9.169	85	0.33	025.42	0
025	Iron (ppm)	0098	200.8	9.200	193.8	18.44	9.169	85	0.38	025.43	0
025	Iron (ppm)	0682	201.2	0.0000	193.8	18.44	9.169	85	0.40	025.43	0
025	Iron (ppm)	0884	201.7	13.30	193.8	18.44	9.169	85	0.42	025.31	0
025	Iron (ppm)	0870	202.5	5.300	193.8	18.44	9.169	85	0.47	025.42	0
025	Iron (ppm)	0914	202.5	17.00	193.8	18.44	9.169	85	0.47	025.43	0
025	Iron (ppm)	0353	204.0	2.000	193.8	18.44	9.169	85	0.55	025.43	0
025	Iron (ppm)	2124	205.6	3.000	193.8	18.44	9.169	85	0.64	025.43	0
025	Iron (ppm)	0689	207.1	20.70	193.8	18.44	9.169	85	0.72	025.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
025	Iron (ppm)	2146	207.7	5.090	193.8	18.44	9.169	85	0.75	025.51	0
025	Iron (ppm)	0598	208.4	3.800	193.8	18.44	9.169	85	0.79	025.41	0
025	Iron (ppm)	2188	209.5	11.00	193.8	18.44	9.169	85	0.85	025.31	0
025	Iron (ppm)	0190	210.6	0.3300	193.8	18.44	9.169	85	0.91	025.42	0
025	Iron (ppm)	2053	213.0	16.00	193.8	18.44	9.169	85	1.04	025.42	0
025	Iron (ppm)	2207	213.0	26.00	193.8	18.44	9.169	85	1.04	025.42	0
025	Iron (ppm)	0035	214.0	6.000	193.8	18.44	9.169	85	1.09	025.42	0
025	Iron (ppm)	2115	214.3	10.35	193.8	18.44	9.169	85	1.11	025.31	0
025	Iron (ppm)	2192	214.5	19.34	193.8	18.44	9.169	85	1.12	025.43	0
025	Iron (ppm)	0066	216.7	4.537	193.8	18.44	9.169	85	1.24	025.41	0
025	Iron (ppm)	0032	218.6	30.00	193.8	18.44	9.169	85	1.34	025.43	0
025	Iron (ppm)	0964	218.8	24.32	193.8	18.44	9.169	85	1.35	025.43	0
025	Iron (ppm)	0520	219.0	34.00	193.8	18.44	9.169	85	1.36	025.41	0
025	Iron (ppm)	2234	222.5	15.96	193.8	18.44	9.169	85	1.56	025.43	0
025	Iron (ppm)	0035	228.5	11.00	193.8	18.44	9.169	85	1.88	025.43	0
025	Iron (ppm)	0038	235.5	20.40	193.8	18.44	9.169	85	2.26	025.31	0
025	Iron (ppm)	0202	236.1	31.90	193.8	18.44	9.169	85	2.29	025.43	0
025	Iron (ppm)	0175	242.0	28.00	193.8	18.44	9.169	85	2.61	025.31	0
025	Iron (ppm)	2089	265.8	12.13	193.8	18.44	9.169	85	3.90	025.43	0
025	Iron (ppm)	0405	296.1	0.3200	193.8	18.44	9.169	85	5.55	025.41	0
025	Iron (ppm)	0003	284.0	230.0	193.8	18.44	9.169	85	4.89	025.41	1
025	Iron (ppm)	0536	469.1	8.960	193.8	18.44	9.169	85	14.93	025.31	2
027	Magnesium (%)	0685	0.0850	0.0100	0.1296	0.0079	0.0045	82	-5.64	027.42	0
027	Magnesium (%)	2187	0.1000	0.0000	0.1296	0.0079	0.0045	82	-3.74	027.43	0
027	Magnesium (%)	0169	0.1050	0.0100	0.1296	0.0079	0.0045	82	-3.11	027.32	0
027	Magnesium (%)	0294	0.1150	0.0100	0.1296	0.0079	0.0045	82	-1.85	027.42	0
027	Magnesium (%)	2129	0.1153	0.0014	0.1296	0.0079	0.0045	82	-1.81	027.42	0
027	Magnesium (%)	0948	0.1170	0.0040	0.1296	0.0079	0.0045	82	-1.59	027.33	0
027	Magnesium (%)	2232	0.1180	0.0120	0.1296	0.0079	0.0045	82	-1.47	027.41	0
027	Magnesium (%)	0407	0.1185	0.0024	0.1296	0.0079	0.0045	82	-1.40	027.41	0
027	Magnesium (%)	0042	0.1190	0.0100	0.1296	0.0079	0.0045	82	-1.34	027.42	0
027	Magnesium (%)	0868	0.1200	0.0000	0.1296	0.0079	0.0045	82	-1.21	027.31	0
027	Magnesium (%)	0910	0.1200	0.0000	0.1296	0.0079	0.0045	82	-1.21	027.41	0
027	Magnesium (%)	0265	0.1200	0.0000	0.1296	0.0079	0.0045	82	-1.21	027.42	0
027	Magnesium (%)	0083	0.1200	0.0000	0.1296	0.0079	0.0045	82	-1.21	027.43	0
027	Magnesium (%)	0510	0.1200	0.0000	0.1296	0.0079	0.0045	82	-1.21	027.43	0
027	Magnesium (%)	0242	0.1200	0.0000	0.1296	0.0079	0.0045	82	-1.21	027.99	0
027	Magnesium (%)	0407	0.1225	0.0030	0.1296	0.0079	0.0045	82	-0.90	027.43	0
027	Magnesium (%)	0154	0.1232	0.0049	0.1296	0.0079	0.0045	82	-0.82	027.52	0
027	Magnesium (%)	0142	0.1248	0.0039	0.1296	0.0079	0.0045	82	-0.61	027.31	0
027	Magnesium (%)	2055	0.1248	0.0008	0.1296	0.0079	0.0045	82	-0.61	027.53	0
027	Magnesium (%)	0003	0.1250	0.0100	0.1296	0.0079	0.0045	82	-0.58	027.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
027	Magnesium (%)	0878	0.1250	0.0100	0.1296	0.0079	0.0045	82	-0.58	027.41	0
027	Magnesium (%)	0353	0.1250	0.0100	0.1296	0.0079	0.0045	82	-0.58	027.43	0
027	Magnesium (%)	0425	0.1250	0.0100	0.1296	0.0079	0.0045	82	-0.58	027.43	0
027	Magnesium (%)	0345	0.1255	0.0030	0.1296	0.0079	0.0045	82	-0.52	027.43	0
027	Magnesium (%)	0148	0.1265	0.0110	0.1296	0.0079	0.0045	82	-0.39	027.41	0
027	Magnesium (%)	0263	0.1265	0.0010	0.1296	0.0079	0.0045	82	-0.39	027.42	0
027	Magnesium (%)	0968	0.1265	0.0010	0.1296	0.0079	0.0045	82	-0.39	027.43	0
027	Magnesium (%)	2219	0.1265	0.0050	0.1296	0.0079	0.0045	82	-0.39	027.44	0
027	Magnesium (%)	2293	0.1265	0.0070	0.1296	0.0079	0.0045	82	-0.39	027.44	0
027	Magnesium (%)	2252	0.1275	0.0030	0.1296	0.0079	0.0045	82	-0.27	027.44	0
027	Magnesium (%)	0038	0.1282	0.0095	0.1296	0.0079	0.0045	82	-0.19	027.31	0
027	Magnesium (%)	0017	0.1299	0.0031	0.1296	0.0079	0.0045	82	0.03	027.43	0
027	Magnesium (%)	0939	0.1300	0.0000	0.1296	0.0079	0.0045	82	0.05	027.31	0
027	Magnesium (%)	2196	0.1300	0.0000	0.1296	0.0079	0.0045	82	0.05	027.31	0
027	Magnesium (%)	0019	0.1300	0.0200	0.1296	0.0079	0.0045	82	0.05	027.41	0
027	Magnesium (%)	2012	0.1300	0.0000	0.1296	0.0079	0.0045	82	0.05	027.41	0
027	Magnesium (%)	0035	0.1300	0.0000	0.1296	0.0079	0.0045	82	0.05	027.42	0
027	Magnesium (%)	0366	0.1300	0.0000	0.1296	0.0079	0.0045	82	0.05	027.42	0
027	Magnesium (%)	0726	0.1300	0.0000	0.1296	0.0079	0.0045	82	0.05	027.42	0
027	Magnesium (%)	0033	0.1300	0.0000	0.1296	0.0079	0.0045	82	0.05	027.43	0
027	Magnesium (%)	0035	0.1300	0.0000	0.1296	0.0079	0.0045	82	0.05	027.43	0
027	Magnesium (%)	0202	0.1300	0.0080	0.1296	0.0079	0.0045	82	0.05	027.43	0
027	Magnesium (%)	0297	0.1300	0.0200	0.1296	0.0079	0.0045	82	0.05	027.43	0
027	Magnesium (%)	0610	0.1300	0.0000	0.1296	0.0079	0.0045	82	0.05	027.43	0
027	Magnesium (%)	2089	0.1300	0.0000	0.1296	0.0079	0.0045	82	0.05	027.43	0
027	Magnesium (%)	2192	0.1300	0.0000	0.1296	0.0079	0.0045	82	0.05	027.43	0
027	Magnesium (%)	0590	0.1300	0.0000	0.1296	0.0079	0.0045	82	0.05	027.99	0
027	Magnesium (%)	0186	0.1303	0.0038	0.1296	0.0079	0.0045	82	0.09	027.52	0
027	Magnesium (%)	2156	0.1305	0.0010	0.1296	0.0079	0.0045	82	0.11	027.44	0
027	Magnesium (%)	0065	0.1310	0.0000	0.1296	0.0079	0.0045	82	0.17	027.44	0
027	Magnesium (%)	0870	0.1314	0.0004	0.1296	0.0079	0.0045	82	0.23	027.42	0
027	Magnesium (%)	0164	0.1315	0.0010	0.1296	0.0079	0.0045	82	0.24	027.44	0
027	Magnesium (%)	2220	0.1315	0.0030	0.1296	0.0079	0.0045	82	0.24	027.44	0
027	Magnesium (%)	0653	0.1320	0.0000	0.1296	0.0079	0.0045	82	0.30	027.41	0
027	Magnesium (%)	0553	0.1320	0.0040	0.1296	0.0079	0.0045	82	0.30	027.53	0
027	Magnesium (%)	0015	0.1330	0.0040	0.1296	0.0079	0.0045	82	0.43	027.53	0
027	Magnesium (%)	0139	0.1334	0.0007	0.1296	0.0079	0.0045	82	0.47	027.41	0
027	Magnesium (%)	0520	0.1340	0.0060	0.1296	0.0079	0.0045	82	0.55	027.41	0
027	Magnesium (%)	0186	0.1344	0.0016	0.1296	0.0079	0.0045	82	0.60	027.42	0
027	Magnesium (%)	0357	0.1349	0.0028	0.1296	0.0079	0.0045	82	0.67	027.42	0
027	Magnesium (%)	0536	0.1350	0.0100	0.1296	0.0079	0.0045	82	0.68	027.31	0
027	Magnesium (%)	0884	0.1350	0.0100	0.1296	0.0079	0.0045	82	0.68	027.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
027	Magnesium (%)	0098	0.1350	0.0100	0.1296	0.0079	0.0045	82	0.68	027.41	0
027	Magnesium (%)	0226	0.1350	0.0100	0.1296	0.0079	0.0045	82	0.68	027.41	0
027	Magnesium (%)	0358	0.1350	0.0100	0.1296	0.0079	0.0045	82	0.68	027.41	0
027	Magnesium (%)	0098	0.1350	0.0100	0.1296	0.0079	0.0045	82	0.68	027.43	0
027	Magnesium (%)	0572	0.1350	0.0000	0.1296	0.0079	0.0045	82	0.68	027.53	0
027	Magnesium (%)	0939	0.1350	0.0100	0.1296	0.0079	0.0045	82	0.68	027.53	0
027	Magnesium (%)	0045	0.1355	0.0090	0.1296	0.0079	0.0045	82	0.74	027.42	0
027	Magnesium (%)	0037	0.1365	0.0030	0.1296	0.0079	0.0045	82	0.87	027.42	0
027	Magnesium (%)	2124	0.1365	0.0010	0.1296	0.0079	0.0045	82	0.87	027.43	0
027	Magnesium (%)	0278	0.1400	0.0000	0.1296	0.0079	0.0045	82	1.31	027.42	0
027	Magnesium (%)	0692	0.1400	0.0000	0.1296	0.0079	0.0045	82	1.31	027.42	0
027	Magnesium (%)	0682	0.1400	0.0000	0.1296	0.0079	0.0045	82	1.31	027.43	0
027	Magnesium (%)	0100	0.1400	0.0000	0.1296	0.0079	0.0045	82	1.31	027.99	0
027	Magnesium (%)	0693	0.1405	0.0090	0.1296	0.0079	0.0045	82	1.38	027.42	0
027	Magnesium (%)	0598	0.1431	0.0005	0.1296	0.0079	0.0045	82	1.70	027.41	0
027	Magnesium (%)	0354	0.1450	0.0100	0.1296	0.0079	0.0045	82	1.94	027.41	0
027	Magnesium (%)	0964	0.1452	0.0295	0.1296	0.0079	0.0045	82	1.96	027.43	0
027	Magnesium (%)	0405	0.1475	0.0010	0.1296	0.0079	0.0045	82	2.26	027.41	0
027	Magnesium (%)	0620	0.1487	0.0034	0.1296	0.0079	0.0045	82	2.41	027.31	0
027	Magnesium (%)	0689	0.2200	0.0000	0.1296	0.0079	0.0045	82	11.42	027.31	0
027	Magnesium (%)	0918	0.1485	0.0430	0.1296	0.0079	0.0045	82	2.39	027.43	1
027	Magnesium (%)	0032	1.300	0.0400	0.1296	0.0079	0.0045	82	147.88	027.43	2
028	Manganese (ppm)	0685	28.40	0.0100	51.60	3.563	1.538	89	-6.51	028.42	0
028	Manganese (ppm)	0689	33.70	1.400	51.60	3.563	1.538	89	-5.02	028.31	0
028	Manganese (ppm)	0620	41.16	2.319	51.60	3.563	1.538	89	-2.93	028.31	0
028	Manganese (ppm)	0017	41.27	0.4529	51.60	3.563	1.538	89	-2.90	028.43	0
028	Manganese (ppm)	0425	45.20	0.1400	51.60	3.563	1.538	89	-1.80	028.43	0
028	Manganese (ppm)	0653	45.97	0.3500	51.60	3.563	1.538	89	-1.58	028.41	0
028	Manganese (ppm)	0870	46.03	0.2100	51.60	3.563	1.538	89	-1.57	028.42	0
028	Manganese (ppm)	0169	46.20	2.000	51.60	3.563	1.538	89	-1.52	028.43	0
028	Manganese (ppm)	2055	46.95	3.100	51.60	3.563	1.538	89	-1.31	028.53	0
028	Manganese (ppm)	0066	47.02	0.8920	51.60	3.563	1.538	89	-1.29	028.44	0
028	Manganese (ppm)	0868	47.05	0.4800	51.60	3.563	1.538	89	-1.28	028.31	0
028	Manganese (ppm)	0001	47.62	3.660	51.60	3.563	1.538	89	-1.12	028.31	0
028	Manganese (ppm)	2188	47.89	3.770	51.60	3.563	1.538	89	-1.04	028.31	0
028	Manganese (ppm)	0948	47.96	0.4450	51.60	3.563	1.538	89	-1.02	028.33	0
028	Manganese (ppm)	0175	48.00	4.000	51.60	3.563	1.538	89	-1.01	028.31	0
028	Manganese (ppm)	0242	48.00	2.000	51.60	3.563	1.538	89	-1.01	028.99	0
028	Manganese (ppm)	0553	48.55	0.1000	51.60	3.563	1.538	89	-0.86	028.53	0
028	Manganese (ppm)	0939	48.73	0.1900	51.60	3.563	1.538	89	-0.81	028.53	0
028	Manganese (ppm)	2022	49.00	1.335	51.60	3.563	1.538	89	-0.73	028.31	0
028	Manganese (ppm)	0164	49.00	0.0000	51.60	3.563	1.538	89	-0.73	028.44	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
028	Manganese (ppm)	2152	49.12	0.0800	51.60	3.563	1.538	89	-0.70	028.44	0
028	Manganese (ppm)	0510	49.50	1.000	51.60	3.563	1.538	89	-0.59	028.43	0
028	Manganese (ppm)	0914	49.55	0.3000	51.60	3.563	1.538	89	-0.58	028.43	0
028	Manganese (ppm)	0407	49.68	0.5145	51.60	3.563	1.538	89	-0.54	028.41	0
028	Manganese (ppm)	2314	49.75	1.300	51.60	3.563	1.538	89	-0.52	028.43	0
028	Manganese (ppm)	2129	49.81	0.5000	51.60	3.563	1.538	89	-0.50	028.42	0
028	Manganese (ppm)	0939	49.84	1.460	51.60	3.563	1.538	89	-0.49	028.31	0
028	Manganese (ppm)	0529	49.90	0.8000	51.60	3.563	1.538	89	-0.48	028.31	0
028	Manganese (ppm)	0910	50.00	6.000	51.60	3.563	1.538	89	-0.45	028.41	0
028	Manganese (ppm)	0366	50.00	0.0000	51.60	3.563	1.538	89	-0.45	028.42	0
028	Manganese (ppm)	0968	50.00	0.0000	51.60	3.563	1.538	89	-0.45	028.43	0
028	Manganese (ppm)	2012	50.03	1.370	51.60	3.563	1.538	89	-0.44	028.41	0
028	Manganese (ppm)	0918	50.25	0.3200	51.60	3.563	1.538	89	-0.38	028.43	0
028	Manganese (ppm)	0263	50.46	0.1440	51.60	3.563	1.538	89	-0.32	028.42	0
028	Manganese (ppm)	2196	50.50	0.0000	51.60	3.563	1.538	89	-0.31	028.31	0
028	Manganese (ppm)	0009	50.50	3.000	51.60	3.563	1.538	89	-0.31	028.42	0
028	Manganese (ppm)	0693	50.80	1.661	51.60	3.563	1.538	89	-0.22	028.42	0
028	Manganese (ppm)	0353	51.00	0.0000	51.60	3.563	1.538	89	-0.17	028.43	0
028	Manganese (ppm)	0619	51.05	0.3000	51.60	3.563	1.538	89	-0.15	028.41	0
028	Manganese (ppm)	0345	51.05	1.300	51.60	3.563	1.538	89	-0.15	028.43	0
028	Manganese (ppm)	0265	51.10	1.200	51.60	3.563	1.538	89	-0.14	028.42	0
028	Manganese (ppm)	0297	51.50	3.000	51.60	3.563	1.538	89	-0.03	028.43	0
028	Manganese (ppm)	0354	51.55	0.3400	51.60	3.563	1.538	89	-0.01	028.41	0
028	Manganese (ppm)	2053	51.70	0.6000	51.60	3.563	1.538	89	0.03	028.42	0
028	Manganese (ppm)	0572	51.75	0.7000	51.60	3.563	1.538	89	0.04	028.53	0
028	Manganese (ppm)	2114	51.98	1.490	51.60	3.563	1.538	89	0.11	028.44	0
028	Manganese (ppm)	0921	52.00	0.0000	51.60	3.563	1.538	89	0.11	028.31	0
028	Manganese (ppm)	0083	52.00	2.000	51.60	3.563	1.538	89	0.11	028.43	0
028	Manganese (ppm)	0015	52.08	6.010	51.60	3.563	1.538	89	0.13	028.53	0
028	Manganese (ppm)	0590	52.30	1.600	51.60	3.563	1.538	89	0.20	028.99	0
028	Manganese (ppm)	0878	52.50	1.000	51.60	3.563	1.538	89	0.25	028.41	0
028	Manganese (ppm)	0035	52.50	1.000	51.60	3.563	1.538	89	0.25	028.42	0
028	Manganese (ppm)	0278	52.50	3.000	51.60	3.563	1.538	89	0.25	028.42	0
028	Manganese (ppm)	0042	52.55	6.500	51.60	3.563	1.538	89	0.27	028.42	0
028	Manganese (ppm)	0019	52.68	3.970	51.60	3.563	1.538	89	0.30	028.41	0
028	Manganese (ppm)	0520	52.76	2.850	51.60	3.563	1.538	89	0.32	028.41	0
028	Manganese (ppm)	0033	52.80	6.000	51.60	3.563	1.538	89	0.34	028.43	0
028	Manganese (ppm)	0045	52.90	2.400	51.60	3.563	1.538	89	0.36	028.42	0
028	Manganese (ppm)	0692	53.00	2.000	51.60	3.563	1.538	89	0.39	028.42	0
028	Manganese (ppm)	2192	53.18	1.090	51.60	3.563	1.538	89	0.44	028.53	0
028	Manganese (ppm)	0226	53.20	2.050	51.60	3.563	1.538	89	0.45	028.41	0
028	Manganese (ppm)	0186	53.50	1.000	51.60	3.563	1.538	89	0.53	028.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
028	Manganese (ppm)	0035	53.50	1.000	51.60	3.563	1.538	89	0.53	028.43	0
028	Manganese (ppm)	0186	53.50	1.000	51.60	3.563	1.538	89	0.53	028.52	0
028	Manganese (ppm)	0148	53.55	1.500	51.60	3.563	1.538	89	0.55	028.41	0
028	Manganese (ppm)	0358	53.63	4.790	51.60	3.563	1.538	89	0.57	028.41	0
028	Manganese (ppm)	0407	53.83	0.3580	51.60	3.563	1.538	89	0.62	028.43	0
028	Manganese (ppm)	0357	53.98	0.0880	51.60	3.563	1.538	89	0.67	028.42	0
028	Manganese (ppm)	0003	54.00	0.0000	51.60	3.563	1.538	89	0.67	028.41	0
028	Manganese (ppm)	0100	54.00	0.0000	51.60	3.563	1.538	89	0.67	028.99	0
028	Manganese (ppm)	2089	54.05	0.5200	51.60	3.563	1.538	89	0.69	028.43	0
028	Manganese (ppm)	0098	54.13	1.060	51.60	3.563	1.538	89	0.71	028.41	0
028	Manganese (ppm)	0682	54.76	0.0000	51.60	3.563	1.538	89	0.89	028.43	0
028	Manganese (ppm)	0032	54.77	1.890	51.60	3.563	1.538	89	0.89	028.43	0
028	Manganese (ppm)	2115	54.95	0.3700	51.60	3.563	1.538	89	0.94	028.31	0
028	Manganese (ppm)	0190	55.03	0.0900	51.60	3.563	1.538	89	0.96	028.42	0
028	Manganese (ppm)	2124	55.21	2.830	51.60	3.563	1.538	89	1.01	028.43	0
028	Manganese (ppm)	0610	55.30	1.000	51.60	3.563	1.538	89	1.04	028.43	0
028	Manganese (ppm)	0098	55.36	0.3600	51.60	3.563	1.538	89	1.06	028.43	0
028	Manganese (ppm)	0037	56.30	0.4000	51.60	3.563	1.538	89	1.32	028.42	0
028	Manganese (ppm)	0598	56.47	0.0500	51.60	3.563	1.538	89	1.37	028.41	0
028	Manganese (ppm)	0038	57.49	0.4300	51.60	3.563	1.538	89	1.65	028.32	0
028	Manganese (ppm)	2207	57.50	7.000	51.60	3.563	1.538	89	1.66	028.42	0
028	Manganese (ppm)	0294	57.75	0.1000	51.60	3.563	1.538	89	1.73	028.42	0
028	Manganese (ppm)	0726	58.68	0.5700	51.60	3.563	1.538	89	1.99	028.42	0
028	Manganese (ppm)	0964	58.71	5.967	51.60	3.563	1.538	89	1.99	028.43	0
028	Manganese (ppm)	0202	58.85	4.590	51.60	3.563	1.538	89	2.03	028.43	0
028	Manganese (ppm)	2234	62.34	2.090	51.60	3.563	1.538	89	3.01	028.43	0
028	Manganese (ppm)	0405	82.85	2.110	51.60	3.563	1.538	89	8.77	028.41	0
028	Manganese (ppm)	0536	28.03	44.15	51.60	3.563	1.538	89	-6.62	028.31	1
028	Manganese (ppm)	0884	158.0	0.3000	51.60	3.563	1.538	89	29.85	028.31	2
031	Phosphorus (%)	0154	0.9099	0.0069	1.087	0.0533	0.0314	122	-3.33	031.52	0
031	Phosphorus (%)	0910	0.9550	0.0900	1.087	0.0533	0.0314	122	-2.48	031.41	0
031	Phosphorus (%)	2188	0.9650	0.0100	1.087	0.0533	0.0314	122	-2.29	031.01	0
031	Phosphorus (%)	0242	0.9750	0.0500	1.087	0.0533	0.0314	122	-2.10	031.99	0
031	Phosphorus (%)	2055	0.9790	0.0215	1.087	0.0533	0.0314	122	-2.03	031.53	0
031	Phosphorus (%)	0626	0.9888	0.0264	1.087	0.0533	0.0314	122	-1.85	031.01	0
031	Phosphorus (%)	0425	1.005	0.0100	1.087	0.0533	0.0314	122	-1.54	031.43	0
031	Phosphorus (%)	0042	1.006	0.0890	1.087	0.0533	0.0314	122	-1.53	031.42	0
031	Phosphorus (%)	0674	1.010	0.0000	1.087	0.0533	0.0314	122	-1.45	031.01	0
031	Phosphorus (%)	0186	1.017	0.0305	1.087	0.0533	0.0314	122	-1.32	031.42	0
031	Phosphorus (%)	0001	1.017	0.0460	1.087	0.0533	0.0314	122	-1.32	031.43	0
031	Phosphorus (%)	0169	1.020	0.0200	1.087	0.0533	0.0314	122	-1.26	031.03	0
031	Phosphorus (%)	0536	1.025	0.0100	1.087	0.0533	0.0314	122	-1.17	031.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	2114	1.025	0.0300	1.087	0.0533	0.0314	122	-1.17	031.44	0
031	Phosphorus (%)	0100	1.025	0.0100	1.087	0.0533	0.0314	122	-1.17	031.99	0
031	Phosphorus (%)	0297	1.030	0.0600	1.087	0.0533	0.0314	122	-1.07	031.43	0
031	Phosphorus (%)	2187	1.030	0.0200	1.087	0.0533	0.0314	122	-1.07	031.43	0
031	Phosphorus (%)	0620	1.032	0.0148	1.087	0.0533	0.0314	122	-1.04	031.01	0
031	Phosphorus (%)	0123	1.035	0.0500	1.087	0.0533	0.0314	122	-0.98	031.41	0
031	Phosphorus (%)	0692	1.035	0.1700	1.087	0.0533	0.0314	122	-0.98	031.42	0
031	Phosphorus (%)	0186	1.038	0.0539	1.087	0.0533	0.0314	122	-0.92	031.52	0
031	Phosphorus (%)	0884	1.040	0.0200	1.087	0.0533	0.0314	122	-0.89	031.01	0
031	Phosphorus (%)	0676	1.040	0.0200	1.087	0.0533	0.0314	122	-0.89	031.99	0
031	Phosphorus (%)	2232	1.043	0.0820	1.087	0.0533	0.0314	122	-0.83	031.41	0
031	Phosphorus (%)	0689	1.045	0.0300	1.087	0.0533	0.0314	122	-0.79	031.01	0
031	Phosphorus (%)	0175	1.050	0.0400	1.087	0.0533	0.0314	122	-0.70	031.01	0
031	Phosphorus (%)	0354	1.050	0.0200	1.087	0.0533	0.0314	122	-0.70	031.41	0
031	Phosphorus (%)	2219	1.050	0.0200	1.087	0.0533	0.0314	122	-0.70	031.44	0
031	Phosphorus (%)	0065	1.053	0.0050	1.087	0.0533	0.0314	122	-0.65	031.44	0
031	Phosphorus (%)	0018	1.055	0.1310	1.087	0.0533	0.0314	122	-0.61	031.01	0
031	Phosphorus (%)	0263	1.055	0.0100	1.087	0.0533	0.0314	122	-0.60	031.42	0
031	Phosphorus (%)	2129	1.057	0.0280	1.087	0.0533	0.0314	122	-0.57	031.42	0
031	Phosphorus (%)	0009	1.058	0.0330	1.087	0.0533	0.0314	122	-0.56	031.42	0
031	Phosphorus (%)	0653	1.059	0.0030	1.087	0.0533	0.0314	122	-0.54	031.41	0
031	Phosphorus (%)	0142	1.059	0.0210	1.087	0.0533	0.0314	122	-0.52	031.01	0
031	Phosphorus (%)	2022	1.060	0.0000	1.087	0.0533	0.0314	122	-0.51	031.01	0
031	Phosphorus (%)	0035	1.060	0.0200	1.087	0.0533	0.0314	122	-0.51	031.42	0
031	Phosphorus (%)	0914	1.060	0.0000	1.087	0.0533	0.0314	122	-0.51	031.43	0
031	Phosphorus (%)	2156	1.060	0.0000	1.087	0.0533	0.0314	122	-0.51	031.44	0
031	Phosphorus (%)	2293	1.060	0.0400	1.087	0.0533	0.0314	122	-0.51	031.44	0
031	Phosphorus (%)	0939	1.060	0.0200	1.087	0.0533	0.0314	122	-0.51	031.53	0
031	Phosphorus (%)	0948	1.065	0.0050	1.087	0.0533	0.0314	122	-0.43	031.01	0
031	Phosphorus (%)	2192	1.065	0.0100	1.087	0.0533	0.0314	122	-0.42	031.01	0
031	Phosphorus (%)	0019	1.065	0.0900	1.087	0.0533	0.0314	122	-0.42	031.41	0
031	Phosphorus (%)	2252	1.065	0.0300	1.087	0.0533	0.0314	122	-0.42	031.44	0
031	Phosphorus (%)	0139	1.066	0.0010	1.087	0.0533	0.0314	122	-0.41	031.41	0
031	Phosphorus (%)	0407	1.068	0.0067	1.087	0.0533	0.0314	122	-0.36	031.41	0
031	Phosphorus (%)	0590	1.070	0.0400	1.087	0.0533	0.0314	122	-0.32	031.01	0
031	Phosphorus (%)	0298	1.070	0.0200	1.087	0.0533	0.0314	122	-0.32	031.41	0
031	Phosphorus (%)	0033	1.070	0.0000	1.087	0.0533	0.0314	122	-0.32	031.43	0
031	Phosphorus (%)	2009	1.072	0.0017	1.087	0.0533	0.0314	122	-0.29	031.01	0
031	Phosphorus (%)	2193	1.075	0.0100	1.087	0.0533	0.0314	122	-0.23	031.01	0
031	Phosphorus (%)	0083	1.075	0.0100	1.087	0.0533	0.0314	122	-0.23	031.43	0
031	Phosphorus (%)	0897	1.080	0.0000	1.087	0.0533	0.0314	122	-0.13	031.01	0
031	Phosphorus (%)	0685	1.080	0.0200	1.087	0.0533	0.0314	122	-0.13	031.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	0357	1.080	0.0012	1.087	0.0533	0.0314	122	-0.13	031.42	0
031	Phosphorus (%)	2295	1.085	0.0100	1.087	0.0533	0.0314	122	-0.04	031.00	0
031	Phosphorus (%)	2115	1.085	0.0100	1.087	0.0533	0.0314	122	-0.04	031.01	0
031	Phosphorus (%)	2196	1.085	0.0000	1.087	0.0533	0.0314	122	-0.04	031.01	0
031	Phosphorus (%)	0144	1.085	0.0300	1.087	0.0533	0.0314	122	-0.04	031.41	0
031	Phosphorus (%)	0010	1.085	0.0300	1.087	0.0533	0.0314	122	-0.04	031.43	0
031	Phosphorus (%)	0918	1.088	0.0010	1.087	0.0533	0.0314	122	0.01	031.43	0
031	Phosphorus (%)	0015	1.090	0.0800	1.087	0.0533	0.0314	122	0.05	031.53	0
031	Phosphorus (%)	0683	1.090	0.0000	1.087	0.0533	0.0314	122	0.05	031.01	0
031	Phosphorus (%)	2181	1.090	0.0000	1.087	0.0533	0.0314	122	0.05	031.01	0
031	Phosphorus (%)	0226	1.090	0.0200	1.087	0.0533	0.0314	122	0.05	031.41	0
031	Phosphorus (%)	0848	1.090	0.0000	1.087	0.0533	0.0314	122	0.05	031.41	0
031	Phosphorus (%)	0294	1.090	0.0200	1.087	0.0533	0.0314	122	0.05	031.42	0
031	Phosphorus (%)	2259	1.094	0.0410	1.087	0.0533	0.0314	122	0.12	031.41	0
031	Phosphorus (%)	0045	1.095	0.1100	1.087	0.0533	0.0314	122	0.15	031.42	0
031	Phosphorus (%)	0366	1.095	0.0300	1.087	0.0533	0.0314	122	0.15	031.42	0
031	Phosphorus (%)	2220	1.095	0.0100	1.087	0.0533	0.0314	122	0.15	031.44	0
031	Phosphorus (%)	2006	1.097	0.0090	1.087	0.0533	0.0314	122	0.17	031.01	0
031	Phosphorus (%)	2109	1.100	0.0250	1.087	0.0533	0.0314	122	0.23	031.01	0
031	Phosphorus (%)	0529	1.100	0.0200	1.087	0.0533	0.0314	122	0.24	031.01	0
031	Phosphorus (%)	0202	1.100	0.0200	1.087	0.0533	0.0314	122	0.24	031.43	0
031	Phosphorus (%)	0682	1.100	0.0000	1.087	0.0533	0.0314	122	0.24	031.43	0
031	Phosphorus (%)	2314	1.100	0.0000	1.087	0.0533	0.0314	122	0.24	031.43	0
031	Phosphorus (%)	0164	1.100	0.0000	1.087	0.0533	0.0314	122	0.24	031.44	0
031	Phosphorus (%)	0051	1.102	0.0182	1.087	0.0533	0.0314	122	0.27	031.41	0
031	Phosphorus (%)	0968	1.103	0.0030	1.087	0.0533	0.0314	122	0.29	031.43	0
031	Phosphorus (%)	0868	1.105	0.0100	1.087	0.0533	0.0314	122	0.33	031.01	0
031	Phosphorus (%)	0036	1.109	0.0046	1.087	0.0533	0.0314	122	0.40	031.43	0
031	Phosphorus (%)	2053	1.110	0.0400	1.087	0.0533	0.0314	122	0.43	031.02	0
031	Phosphorus (%)	0939	1.110	0.0200	1.087	0.0533	0.0314	122	0.43	031.01	0
031	Phosphorus (%)	0001	1.111	0.0320	1.087	0.0533	0.0314	122	0.45	031.03	0
031	Phosphorus (%)	0405	1.111	0.0040	1.087	0.0533	0.0314	122	0.45	031.41	0
031	Phosphorus (%)	0036	1.111	0.0016	1.087	0.0533	0.0314	122	0.45	031.03	0
031	Phosphorus (%)	0265	1.115	0.0700	1.087	0.0533	0.0314	122	0.52	031.42	0
031	Phosphorus (%)	0038	1.118	0.0275	1.087	0.0533	0.0314	122	0.57	031.01	0
031	Phosphorus (%)	0017	1.119	0.0997	1.087	0.0533	0.0314	122	0.60	031.43	0
031	Phosphorus (%)	0098	1.123	0.0610	1.087	0.0533	0.0314	122	0.66	031.41	0
031	Phosphorus (%)	0278	1.125	0.0500	1.087	0.0533	0.0314	122	0.71	031.42	0
031	Phosphorus (%)	0610	1.125	0.0100	1.087	0.0533	0.0314	122	0.71	031.43	0
031	Phosphorus (%)	0510	1.130	0.0000	1.087	0.0533	0.0314	122	0.80	031.43	0
031	Phosphorus (%)	0553	1.130	0.0800	1.087	0.0533	0.0314	122	0.80	031.53	0
031	Phosphorus (%)	0098	1.132	0.0170	1.087	0.0533	0.0314	122	0.83	031.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	0148	1.135	0.0300	1.087	0.0533	0.0314	122	0.90	031.41	0
031	Phosphorus (%)	0345	1.135	0.0300	1.087	0.0533	0.0314	122	0.90	031.43	0
031	Phosphorus (%)	2089	1.135	0.0300	1.087	0.0533	0.0314	122	0.90	031.43	0
031	Phosphorus (%)	0407	1.137	0.0120	1.087	0.0533	0.0314	122	0.93	031.43	0
031	Phosphorus (%)	0032	1.138	0.1420	1.087	0.0533	0.0314	122	0.95	031.43	0
031	Phosphorus (%)	0870	1.140	0.0790	1.087	0.0533	0.0314	122	0.98	031.42	0
031	Phosphorus (%)	0878	1.140	0.0000	1.087	0.0533	0.0314	122	0.99	031.01	0
031	Phosphorus (%)	2124	1.140	0.0400	1.087	0.0533	0.0314	122	0.99	031.43	0
031	Phosphorus (%)	0619	1.150	0.0600	1.087	0.0533	0.0314	122	1.18	031.41	0
031	Phosphorus (%)	0014	1.160	0.0200	1.087	0.0533	0.0314	122	1.37	031.42	0
031	Phosphorus (%)	0035	1.160	0.0000	1.087	0.0533	0.0314	122	1.37	031.43	0
031	Phosphorus (%)	0358	1.160	0.1400	1.087	0.0533	0.0314	122	1.37	031.41	0
031	Phosphorus (%)	2012	1.160	0.1000	1.087	0.0533	0.0314	122	1.37	031.41	0
031	Phosphorus (%)	0066	1.161	0.0630	1.087	0.0533	0.0314	122	1.38	031.44	0
031	Phosphorus (%)	0726	1.170	0.0000	1.087	0.0533	0.0314	122	1.55	031.42	0
031	Phosphorus (%)	0964	1.173	0.0137	1.087	0.0533	0.0314	122	1.61	031.43	0
031	Phosphorus (%)	0190	1.185	0.0100	1.087	0.0533	0.0314	122	1.84	031.42	0
031	Phosphorus (%)	0037	1.195	0.0090	1.087	0.0533	0.0314	122	2.01	031.42	0
031	Phosphorus (%)	0353	1.230	0.0400	1.087	0.0533	0.0314	122	2.68	031.43	0
031	Phosphorus (%)	0598	1.237	0.0090	1.087	0.0533	0.0314	122	2.80	031.41	0
031	Phosphorus (%)	0572	1.255	0.0100	1.087	0.0533	0.0314	122	3.15	031.53	0
031	Phosphorus (%)	0003	1.260	0.0000	1.087	0.0533	0.0314	122	3.24	031.41	0
031	Phosphorus (%)	2234	1.260	0.0600	1.087	0.0533	0.0314	122	3.24	031.43	0
031	Phosphorus (%)	2207	1.290	0.2000	1.087	0.0533	0.0314	122	3.81	031.42	0
031	Phosphorus (%)	0520	1.423	0.1120	1.087	0.0533	0.0314	122	6.30	031.41	0
031	Phosphorus (%)	0693	0.9195	0.3450	1.087	0.0533	0.0314	122	-3.15	031.42	1
031	Phosphorus (%)	0687	0.9500	0.3000	1.087	0.0533	0.0314	122	-2.57	031.01	1
032	Potassium (%)	0154	0.6010	0.0243	0.7567	0.0440	0.0186	81	-3.54	032.52	0
032	Potassium (%)	0169	0.6200	0.0400	0.7567	0.0440	0.0186	81	-3.11	032.32	0
032	Potassium (%)	0294	0.6450	0.0100	0.7567	0.0440	0.0186	81	-2.54	032.42	0
032	Potassium (%)	0425	0.6650	0.0100	0.7567	0.0440	0.0186	81	-2.09	032.43	0
032	Potassium (%)	0536	0.6700	0.0000	0.7567	0.0440	0.0186	81	-1.97	032.31	0
032	Potassium (%)	2146	0.6700	0.0200	0.7567	0.0440	0.0186	81	-1.97	032.51	0
032	Potassium (%)	0123	0.6800	0.0000	0.7567	0.0440	0.0186	81	-1.75	032.41	0
032	Potassium (%)	0019	0.6900	0.0400	0.7567	0.0440	0.0186	81	-1.52	032.41	0
032	Potassium (%)	0144	0.6900	0.0200	0.7567	0.0440	0.0186	81	-1.52	032.41	0
032	Potassium (%)	2232	0.6955	0.0630	0.7567	0.0440	0.0186	81	-1.39	032.41	0
032	Potassium (%)	0042	0.6965	0.1030	0.7567	0.0440	0.0186	81	-1.37	032.42	0
032	Potassium (%)	2129	0.6971	0.0049	0.7567	0.0440	0.0186	81	-1.36	032.42	0
032	Potassium (%)	0683	0.7000	0.0200	0.7567	0.0440	0.0186	81	-1.29	032.31	0
032	Potassium (%)	0918	0.7100	0.0180	0.7567	0.0440	0.0186	81	-1.06	032.43	0
032	Potassium (%)	0003	0.7200	0.1200	0.7567	0.0440	0.0186	81	-0.84	032.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
032	Potassium (%)	0265	0.7200	0.0200	0.7567	0.0440	0.0186	81	-0.84	032.42	0
032	Potassium (%)	0242	0.7200	0.0200	0.7567	0.0440	0.0186	81	-0.84	032.99	0
032	Potassium (%)	0083	0.7250	0.0100	0.7567	0.0440	0.0186	81	-0.72	032.43	0
032	Potassium (%)	2055	0.7288	0.0288	0.7567	0.0440	0.0186	81	-0.64	032.53	0
032	Potassium (%)	0910	0.7300	0.0200	0.7567	0.0440	0.0186	81	-0.61	032.41	0
032	Potassium (%)	0868	0.7350	0.0100	0.7567	0.0440	0.0186	81	-0.49	032.31	0
032	Potassium (%)	2187	0.7350	0.0100	0.7567	0.0440	0.0186	81	-0.49	032.43	0
032	Potassium (%)	2219	0.7350	0.0300	0.7567	0.0440	0.0186	81	-0.49	032.44	0
032	Potassium (%)	0100	0.7350	0.0300	0.7567	0.0440	0.0186	81	-0.49	032.99	0
032	Potassium (%)	2012	0.7400	0.0000	0.7567	0.0440	0.0186	81	-0.38	032.41	0
032	Potassium (%)	0682	0.7400	0.0000	0.7567	0.0440	0.0186	81	-0.38	032.43	0
032	Potassium (%)	2220	0.7400	0.0200	0.7567	0.0440	0.0186	81	-0.38	032.44	0
032	Potassium (%)	2252	0.7400	0.0200	0.7567	0.0440	0.0186	81	-0.38	032.44	0
032	Potassium (%)	0590	0.7400	0.0200	0.7567	0.0440	0.0186	81	-0.38	032.99	0
032	Potassium (%)	0354	0.7450	0.0100	0.7567	0.0440	0.0186	81	-0.27	032.41	0
032	Potassium (%)	0345	0.7455	0.0030	0.7567	0.0440	0.0186	81	-0.26	032.43	0
032	Potassium (%)	2156	0.7500	0.0000	0.7567	0.0440	0.0186	81	-0.15	032.44	0
032	Potassium (%)	0186	0.7504	0.0049	0.7567	0.0440	0.0186	81	-0.15	032.42	0
032	Potassium (%)	0948	0.7540	0.0060	0.7567	0.0440	0.0186	81	-0.06	032.31	0
032	Potassium (%)	2293	0.7550	0.0300	0.7567	0.0440	0.0186	81	-0.04	032.44	0
032	Potassium (%)	0139	0.7570	0.0020	0.7567	0.0440	0.0186	81	0.01	032.41	0
032	Potassium (%)	0186	0.7577	0.0070	0.7567	0.0440	0.0186	81	0.02	032.52	0
032	Potassium (%)	0553	0.7595	0.0010	0.7567	0.0440	0.0186	81	0.06	032.53	0
032	Potassium (%)	0035	0.7600	0.0000	0.7567	0.0440	0.0186	81	0.07	032.42	0
032	Potassium (%)	0065	0.7600	0.0000	0.7567	0.0440	0.0186	81	0.07	032.44	0
032	Potassium (%)	0015	0.7600	0.0400	0.7567	0.0440	0.0186	81	0.07	032.53	0
032	Potassium (%)	0148	0.7630	0.0220	0.7567	0.0440	0.0186	81	0.14	032.41	0
032	Potassium (%)	0142	0.7644	0.0260	0.7567	0.0440	0.0186	81	0.17	032.31	0
032	Potassium (%)	0035	0.7650	0.0100	0.7567	0.0440	0.0186	81	0.19	032.43	0
032	Potassium (%)	0297	0.7650	0.0100	0.7567	0.0440	0.0186	81	0.19	032.43	0
032	Potassium (%)	0939	0.7650	0.0100	0.7567	0.0440	0.0186	81	0.19	032.53	0
032	Potassium (%)	0693	0.7665	0.0390	0.7567	0.0440	0.0186	81	0.22	032.42	0
032	Potassium (%)	0164	0.7685	0.0090	0.7567	0.0440	0.0186	81	0.27	032.44	0
032	Potassium (%)	2053	0.7700	0.0000	0.7567	0.0440	0.0186	81	0.30	032.42	0
032	Potassium (%)	0033	0.7700	0.0200	0.7567	0.0440	0.0186	81	0.30	032.43	0
032	Potassium (%)	0098	0.7700	0.0200	0.7567	0.0440	0.0186	81	0.30	032.43	0
032	Potassium (%)	0045	0.7730	0.0100	0.7567	0.0440	0.0186	81	0.37	032.42	0
032	Potassium (%)	0653	0.7745	0.0050	0.7567	0.0440	0.0186	81	0.40	032.41	0
032	Potassium (%)	0939	0.7750	0.0100	0.7567	0.0440	0.0186	81	0.42	032.31	0
032	Potassium (%)	0510	0.7750	0.0100	0.7567	0.0440	0.0186	81	0.42	032.43	0
032	Potassium (%)	0870	0.7759	0.0049	0.7567	0.0440	0.0186	81	0.43	032.42	0
032	Potassium (%)	0405	0.7765	0.0050	0.7567	0.0440	0.0186	81	0.45	032.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
032	Potassium (%)	0263	0.7780	0.0040	0.7567	0.0440	0.0186	81	0.48	032.42	0
032	Potassium (%)	2089	0.7800	0.0400	0.7567	0.0440	0.0186	81	0.53	032.43	0
032	Potassium (%)	0407	0.7840	0.0000	0.7567	0.0440	0.0186	81	0.62	032.43	0
032	Potassium (%)	0226	0.7850	0.0100	0.7567	0.0440	0.0186	81	0.64	032.41	0
032	Potassium (%)	0878	0.7850	0.0100	0.7567	0.0440	0.0186	81	0.64	032.41	0
032	Potassium (%)	0357	0.7867	0.0517	0.7567	0.0440	0.0186	81	0.68	032.42	0
032	Potassium (%)	0358	0.7900	0.0400	0.7567	0.0440	0.0186	81	0.76	032.41	0
032	Potassium (%)	0407	0.7956	0.0010	0.7567	0.0440	0.0186	81	0.88	032.41	0
032	Potassium (%)	2124	0.7970	0.0000	0.7567	0.0440	0.0186	81	0.92	032.43	0
032	Potassium (%)	0202	0.7975	0.0230	0.7567	0.0440	0.0186	81	0.93	032.43	0
032	Potassium (%)	0098	0.8000	0.0000	0.7567	0.0440	0.0186	81	0.98	032.41	0
032	Potassium (%)	0037	0.8000	0.0000	0.7567	0.0440	0.0186	81	0.98	032.42	0
032	Potassium (%)	0278	0.8000	0.0200	0.7567	0.0440	0.0186	81	0.98	032.42	0
032	Potassium (%)	0884	0.8050	0.0100	0.7567	0.0440	0.0186	81	1.10	032.02	0
032	Potassium (%)	0692	0.8050	0.0100	0.7567	0.0440	0.0186	81	1.10	032.42	0
032	Potassium (%)	0968	0.8070	0.0040	0.7567	0.0440	0.0186	81	1.14	032.43	0
032	Potassium (%)	0017	0.8078	0.0355	0.7567	0.0440	0.0186	81	1.16	032.43	0
032	Potassium (%)	0598	0.8090	0.0017	0.7567	0.0440	0.0186	81	1.19	032.41	0
032	Potassium (%)	0726	0.8100	0.0000	0.7567	0.0440	0.0186	81	1.21	032.42	0
032	Potassium (%)	0610	0.8150	0.0100	0.7567	0.0440	0.0186	81	1.32	032.43	0
032	Potassium (%)	0353	0.8200	0.0400	0.7567	0.0440	0.0186	81	1.44	032.43	0
032	Potassium (%)	0520	0.8225	0.0210	0.7567	0.0440	0.0186	81	1.50	032.41	0
032	Potassium (%)	0572	0.8535	0.0650	0.7567	0.0440	0.0186	81	2.20	032.53	0
032	Potassium (%)	0366	0.8600	0.0600	0.7567	0.0440	0.0186	81	2.35	032.42	0
032	Potassium (%)	0964	0.8230	0.1664	0.7567	0.0440	0.0186	81	1.51	032.43	1
032	Potassium (%)	0032	7.637	0.2830	0.7567	0.0440	0.0186	81	156.43	032.43	2
033	Salt as chloride (%)	0144	0.4600	0.1000	1.153	0.0835	0.0205	46	-8.31	033.03	0
033	Salt as chloride (%)	2295	0.9850	0.0100	1.153	0.0835	0.0205	46	-2.02	033.00	0
033	Salt as chloride (%)	2129	1.029	0.0290	1.153	0.0835	0.0205	46	-1.50	033.99	0
033	Salt as chloride (%)	0536	1.035	0.0100	1.153	0.0835	0.0205	46	-1.42	033.99	0
033	Salt as chloride (%)	0539	1.040	0.0000	1.153	0.0835	0.0205	46	-1.36	033.00	0
033	Salt as chloride (%)	2073	1.055	0.0100	1.153	0.0835	0.0205	46	-1.18	033.99	0
033	Salt as chloride (%)	2043	1.060	0.0200	1.153	0.0835	0.0205	46	-1.12	033.05	0
033	Salt as chloride (%)	0897	1.080	0.0000	1.153	0.0835	0.0205	46	-0.88	033.00	0
033	Salt as chloride (%)	2268	1.080	0.0400	1.153	0.0835	0.0205	46	-0.88	033.01	0
033	Salt as chloride (%)	0265	1.080	0.0200	1.153	0.0835	0.0205	46	-0.88	033.03	0
033	Salt as chloride (%)	0868	1.080	0.0200	1.153	0.0835	0.0205	46	-0.88	033.05	0
033	Salt as chloride (%)	0353	1.095	0.0100	1.153	0.0835	0.0205	46	-0.70	033.00	0
033	Salt as chloride (%)	0619	1.095	0.0100	1.153	0.0835	0.0205	46	-0.70	033.01	0
033	Salt as chloride (%)	0045	1.100	0.0400	1.153	0.0835	0.0205	46	-0.64	033.00	0
033	Salt as chloride (%)	2193	1.100	0.0400	1.153	0.0835	0.0205	46	-0.64	033.00	0
033	Salt as chloride (%)	2196	1.130	0.0000	1.153	0.0835	0.0205	46	-0.28	033.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
033	Salt as chloride (%)	0878	1.140	0.0200	1.153	0.0835	0.0205	46	-0.16	033.00	0
033	Salt as chloride (%)	0910	1.140	0.0200	1.153	0.0835	0.0205	46	-0.16	033.99	0
033	Salt as chloride (%)	0693	1.143	0.0610	1.153	0.0835	0.0205	46	-0.13	033.00	0
033	Salt as chloride (%)	0083	1.150	0.0000	1.153	0.0835	0.0205	46	-0.04	033.01	0
033	Salt as chloride (%)	0510	1.154	0.0000	1.153	0.0835	0.0205	46	0.01	033.01	0
033	Salt as chloride (%)	0689	1.155	0.0100	1.153	0.0835	0.0205	46	0.02	033.05	0
033	Salt as chloride (%)	0309	1.156	0.0279	1.153	0.0835	0.0205	46	0.03	033.00	0
033	Salt as chloride (%)	0148	1.170	0.0200	1.153	0.0835	0.0205	46	0.20	033.01	0
033	Salt as chloride (%)	0098	1.180	0.0200	1.153	0.0835	0.0205	46	0.32	033.01	0
033	Salt as chloride (%)	2109	1.182	0.0090	1.153	0.0835	0.0205	46	0.34	033.01	0
033	Salt as chloride (%)	0884	1.185	0.0100	1.153	0.0835	0.0205	46	0.38	033.00	0
033	Salt as chloride (%)	2022	1.185	0.0300	1.153	0.0835	0.0205	46	0.38	033.00	0
033	Salt as chloride (%)	0590	1.185	0.0500	1.153	0.0835	0.0205	46	0.38	033.01	0
033	Salt as chloride (%)	0683	1.185	0.0100	1.153	0.0835	0.0205	46	0.38	033.01	0
033	Salt as chloride (%)	0407	1.188	0.0060	1.153	0.0835	0.0205	46	0.41	033.01	0
033	Salt as chloride (%)	2076	1.190	0.0000	1.153	0.0835	0.0205	46	0.44	033.00	0
033	Salt as chloride (%)	0242	1.190	0.0000	1.153	0.0835	0.0205	46	0.44	033.01	0
033	Salt as chloride (%)	0354	1.195	0.0100	1.153	0.0835	0.0205	46	0.50	033.01	0
033	Salt as chloride (%)	0100	1.200	0.0000	1.153	0.0835	0.0205	46	0.56	033.01	0
033	Salt as chloride (%)	0164	1.200	0.0000	1.153	0.0835	0.0205	46	0.56	033.01	0
033	Salt as chloride (%)	0948	1.200	0.0000	1.153	0.0835	0.0205	46	0.56	033.01	0
033	Salt as chloride (%)	0682	1.200	0.0000	1.153	0.0835	0.0205	46	0.56	033.99	0
033	Salt as chloride (%)	2181	1.200	0.0000	1.153	0.0835	0.0205	46	0.56	033.99	0
033	Salt as chloride (%)	0226	1.220	0.0400	1.153	0.0835	0.0205	46	0.80	033.01	0
033	Salt as chloride (%)	0298	1.260	0.0200	1.153	0.0835	0.0205	46	1.28	033.00	0
033	Salt as chloride (%)	0186	1.297	0.0310	1.153	0.0835	0.0205	46	1.71	033.01	0
033	Salt as chloride (%)	0032	1.499	0.0275	1.153	0.0835	0.0205	46	4.14	033.99	0
033	Salt as chloride (%)	0278	1.510	0.0200	1.153	0.0835	0.0205	46	4.27	033.01	0
033	Salt as chloride (%)	0017	1.520	0.0529	1.153	0.0835	0.0205	46	4.39	033.99	0
033	Salt as chloride (%)	0358	1.615	0.0900	1.153	0.0835	0.0205	46	5.53	033.99	0
033	Salt as chloride (%)	0366	1.310	0.1600	1.153	0.0835	0.0205	46	1.87	033.00	1
034	Selenium (ppm)	2234	0.1800	0.0200	0.6850	0.2648	0.0469	25	-1.91	034.43	0
034	Selenium (ppm)	2114	0.4300	0.0200	0.6850	0.2648	0.0469	25	-0.96	034.43	0
034	Selenium (ppm)	0964	0.4379	0.0092	0.6850	0.2648	0.0469	25	-0.93	034.43	0
034	Selenium (ppm)	0619	0.4830	0.0020	0.6850	0.2648	0.0469	25	-0.76	034.41	0
034	Selenium (ppm)	0045	0.5170	0.0680	0.6850	0.2648	0.0469	25	-0.63	034.04	0
034	Selenium (ppm)	0148	0.5435	0.0070	0.6850	0.2648	0.0469	25	-0.53	034.41	0
034	Selenium (ppm)	2055	0.5550	0.0100	0.6850	0.2648	0.0469	25	-0.49	034.53	0
034	Selenium (ppm)	0186	0.5650	0.0198	0.6850	0.2648	0.0469	25	-0.45	034.52	0
034	Selenium (ppm)	0009	0.5719	0.1040	0.6850	0.2648	0.0469	25	-0.43	034.53	0
034	Selenium (ppm)	0164	0.5725	0.0070	0.6850	0.2648	0.0469	25	-0.42	034.53	0
034	Selenium (ppm)	0553	0.6095	0.0150	0.6850	0.2648	0.0469	25	-0.28	034.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
034	Selenium (ppm)	0169	0.6150	0.0300	0.6850	0.2648	0.0469	25	-0.26	034.04	0
034	Selenium (ppm)	0610	0.6150	0.0300	0.6850	0.2648	0.0469	25	-0.26	034.53	0
034	Selenium (ppm)	0910	0.6200	0.0200	0.6850	0.2648	0.0469	25	-0.25	034.52	0
034	Selenium (ppm)	0032	0.6305	0.0350	0.6850	0.2648	0.0469	25	-0.21	034.53	0
034	Selenium (ppm)	0015	0.6350	0.1500	0.6850	0.2648	0.0469	25	-0.19	034.53	0
034	Selenium (ppm)	0038	0.6428	0.0084	0.6850	0.2648	0.0469	25	-0.16	034.01	0
034	Selenium (ppm)	2207	0.7000	0.2000	0.6850	0.2648	0.0469	25	0.06	034.52	0
034	Selenium (ppm)	2012	0.7550	0.0440	0.6850	0.2648	0.0469	25	0.26	034.53	0
034	Selenium (ppm)	0918	0.9265	0.0412	0.6850	0.2648	0.0469	25	0.91	034.53	0
034	Selenium (ppm)	0968	1.090	0.0530	0.6850	0.2648	0.0469	25	1.53	034.43	0
034	Selenium (ppm)	0407	1.680	0.0800	0.6850	0.2648	0.0469	25	3.76	034.41	0
034	Selenium (ppm)	0407	1.740	0.1200	0.6850	0.2648	0.0469	25	3.98	034.43	0
034	Selenium (ppm)	0190	1.750	0.0800	0.6850	0.2648	0.0469	25	4.02	034.99	0
034	Selenium (ppm)	0682	2.200	0.0000	0.6850	0.2648	0.0469	25	5.72	034.43	0
034	Selenium (ppm)	0939	0.9100	0.3600	0.6850	0.2648	0.0469	25	0.85	034.53	1
034	Selenium (ppm)	0692	< 1		0.6850	0.2648	0.0469	25		034.42	5
034	Selenium (ppm)	2124	< 1		0.6850	0.2648	0.0469	25		034.43	5
035	Sodium (%)	0169	0.4750	0.0700	0.5808	0.0351	0.0171	95	-3.02	035.32	0
035	Sodium (%)	0366	0.4765	0.0550	0.5808	0.0351	0.0171	95	-2.98	035.42	0
035	Sodium (%)	0175	0.4800	0.0400	0.5808	0.0351	0.0171	95	-2.88	035.31	0
035	Sodium (%)	0683	0.5000	0.0000	0.5808	0.0351	0.0171	95	-2.31	035.31	0
035	Sodium (%)	2259	0.5065	0.0290	0.5808	0.0351	0.0171	95	-2.12	035.41	0
035	Sodium (%)	0042	0.5090	0.0120	0.5808	0.0351	0.0171	95	-2.05	035.42	0
035	Sodium (%)	2114	0.5215	0.0010	0.5808	0.0351	0.0171	95	-1.69	035.41	0
035	Sodium (%)	0425	0.5250	0.0100	0.5808	0.0351	0.0171	95	-1.59	035.43	0
035	Sodium (%)	0294	0.5300	0.0000	0.5808	0.0351	0.0171	95	-1.45	035.42	0
035	Sodium (%)	0036	0.5325	0.0248	0.5808	0.0351	0.0171	95	-1.38	035.43	0
035	Sodium (%)	2129	0.5352	0.0001	0.5808	0.0351	0.0171	95	-1.30	035.42	0
035	Sodium (%)	0154	0.5360	0.0047	0.5808	0.0351	0.0171	95	-1.28	035.52	0
035	Sodium (%)	0144	0.5400	0.0200	0.5808	0.0351	0.0171	95	-1.16	035.41	0
035	Sodium (%)	0910	0.5400	0.0000	0.5808	0.0351	0.0171	95	-1.16	035.41	0
035	Sodium (%)	0265	0.5400	0.0200	0.5808	0.0351	0.0171	95	-1.16	035.42	0
035	Sodium (%)	0242	0.5400	0.0000	0.5808	0.0351	0.0171	95	-1.16	035.99	0
035	Sodium (%)	0598	0.5448	0.0052	0.5808	0.0351	0.0171	95	-1.03	035.41	0
035	Sodium (%)	2232	0.5455	0.0250	0.5808	0.0351	0.0171	95	-1.01	035.41	0
035	Sodium (%)	0918	0.5460	0.0220	0.5808	0.0351	0.0171	95	-0.99	035.43	0
035	Sodium (%)	0298	0.5500	0.0200	0.5808	0.0351	0.0171	95	-0.88	035.41	0
035	Sodium (%)	0914	0.5505	0.0530	0.5808	0.0351	0.0171	95	-0.87	035.43	0
035	Sodium (%)	2115	0.5551	0.0170	0.5808	0.0351	0.0171	95	-0.73	035.31	0
035	Sodium (%)	0148	0.5565	0.0010	0.5808	0.0351	0.0171	95	-0.69	035.41	0
035	Sodium (%)	0619	0.5575	0.0230	0.5808	0.0351	0.0171	95	-0.67	035.41	0
035	Sodium (%)	0354	0.5600	0.0000	0.5808	0.0351	0.0171	95	-0.59	035.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
035	Sodium (%)	2219	0.5600	0.0200	0.5808	0.0351	0.0171	95	-0.59	035.41	0
035	Sodium (%)	0037	0.5620	0.0060	0.5808	0.0351	0.0171	95	-0.54	035.42	0
035	Sodium (%)	0884	0.5650	0.0100	0.5808	0.0351	0.0171	95	-0.45	035.02	0
035	Sodium (%)	2187	0.5650	0.0100	0.5808	0.0351	0.0171	95	-0.45	035.43	0
035	Sodium (%)	0529	0.5660	0.0060	0.5808	0.0351	0.0171	95	-0.42	035.31	0
035	Sodium (%)	0345	0.5675	0.0110	0.5808	0.0351	0.0171	95	-0.38	035.43	0
035	Sodium (%)	0142	0.5687	0.0148	0.5808	0.0351	0.0171	95	-0.35	035.31	0
035	Sodium (%)	0065	0.5700	0.0000	0.5808	0.0351	0.0171	95	-0.31	035.41	0
035	Sodium (%)	2293	0.5700	0.0400	0.5808	0.0351	0.0171	95	-0.31	035.41	0
035	Sodium (%)	0035	0.5700	0.0200	0.5808	0.0351	0.0171	95	-0.31	035.42	0
035	Sodium (%)	0939	0.5700	0.0200	0.5808	0.0351	0.0171	95	-0.31	035.53	0
035	Sodium (%)	0536	0.5750	0.0300	0.5808	0.0351	0.0171	95	-0.17	035.31	0
035	Sodium (%)	2156	0.5750	0.0100	0.5808	0.0351	0.0171	95	-0.17	035.41	0
035	Sodium (%)	2252	0.5750	0.0100	0.5808	0.0351	0.0171	95	-0.17	035.41	0
035	Sodium (%)	2196	0.5800	0.0000	0.5808	0.0351	0.0171	95	-0.02	035.31	0
035	Sodium (%)	0682	0.5800	0.0000	0.5808	0.0351	0.0171	95	-0.02	035.43	0
035	Sodium (%)	0051	0.5813	0.0152	0.5808	0.0351	0.0171	95	0.01	035.41	0
035	Sodium (%)	0553	0.5840	0.0180	0.5808	0.0351	0.0171	95	0.09	035.53	0
035	Sodium (%)	0186	0.5850	0.0129	0.5808	0.0351	0.0171	95	0.12	035.52	0
035	Sodium (%)	0019	0.5850	0.0500	0.5808	0.0351	0.0171	95	0.12	035.41	0
035	Sodium (%)	0123	0.5850	0.0100	0.5808	0.0351	0.0171	95	0.12	035.41	0
035	Sodium (%)	0653	0.5850	0.0020	0.5808	0.0351	0.0171	95	0.12	035.41	0
035	Sodium (%)	2012	0.5850	0.0100	0.5808	0.0351	0.0171	95	0.12	035.41	0
035	Sodium (%)	0278	0.5850	0.0300	0.5808	0.0351	0.0171	95	0.12	035.42	0
035	Sodium (%)	0083	0.5850	0.0100	0.5808	0.0351	0.0171	95	0.12	035.43	0
035	Sodium (%)	0015	0.5865	0.0210	0.5808	0.0351	0.0171	95	0.16	035.53	0
035	Sodium (%)	2055	0.5867	0.0086	0.5808	0.0351	0.0171	95	0.17	035.53	0
035	Sodium (%)	0870	0.5897	0.0067	0.5808	0.0351	0.0171	95	0.25	035.42	0
035	Sodium (%)	0939	0.5900	0.0000	0.5808	0.0351	0.0171	95	0.26	035.31	0
035	Sodium (%)	2220	0.5900	0.0400	0.5808	0.0351	0.0171	95	0.26	035.41	0
035	Sodium (%)	0035	0.5900	0.0000	0.5808	0.0351	0.0171	95	0.26	035.43	0
035	Sodium (%)	0098	0.5900	0.0200	0.5808	0.0351	0.0171	95	0.26	035.43	0
035	Sodium (%)	0003	0.5900	0.0600	0.5808	0.0351	0.0171	95	0.26	035.41	0
035	Sodium (%)	0297	0.5900	0.0600	0.5808	0.0351	0.0171	95	0.26	035.43	0
035	Sodium (%)	0407	0.5905	0.0050	0.5808	0.0351	0.0171	95	0.28	035.43	0
035	Sodium (%)	0407	0.5926	0.0036	0.5808	0.0351	0.0171	95	0.34	035.41	0
035	Sodium (%)	2089	0.5950	0.0100	0.5808	0.0351	0.0171	95	0.40	035.43	0
035	Sodium (%)	0100	0.5950	0.0100	0.5808	0.0351	0.0171	95	0.40	035.99	0
035	Sodium (%)	0590	0.5950	0.0100	0.5808	0.0351	0.0171	95	0.40	035.99	0
035	Sodium (%)	0186	0.5971	0.0058	0.5808	0.0351	0.0171	95	0.46	035.42	0
035	Sodium (%)	0263	0.5975	0.0030	0.5808	0.0351	0.0171	95	0.48	035.42	0
035	Sodium (%)	0868	0.5980	0.0080	0.5808	0.0351	0.0171	95	0.49	035.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
035	Sodium (%)	0017	0.5980	0.0208	0.5808	0.0351	0.0171	95	0.49	035.43	0
035	Sodium (%)	0164	0.5985	0.0070	0.5808	0.0351	0.0171	95	0.50	035.41	0
035	Sodium (%)	0868	0.6000	0.0000	0.5808	0.0351	0.0171	95	0.55	035.31	0
035	Sodium (%)	0693	0.6000	0.0280	0.5808	0.0351	0.0171	95	0.55	035.42	0
035	Sodium (%)	0033	0.6000	0.0200	0.5808	0.0351	0.0171	95	0.55	035.43	0
035	Sodium (%)	0572	0.6000	0.0240	0.5808	0.0351	0.0171	95	0.55	035.53	0
035	Sodium (%)	0353	0.6032	0.0279	0.5808	0.0351	0.0171	95	0.64	035.43	0
035	Sodium (%)	0878	0.6050	0.0100	0.5808	0.0351	0.0171	95	0.69	035.41	0
035	Sodium (%)	0139	0.6065	0.0010	0.5808	0.0351	0.0171	95	0.73	035.41	0
035	Sodium (%)	0948	0.6095	0.0050	0.5808	0.0351	0.0171	95	0.82	035.31	0
035	Sodium (%)	0098	0.6100	0.0000	0.5808	0.0351	0.0171	95	0.83	035.41	0
035	Sodium (%)	0202	0.6110	0.0360	0.5808	0.0351	0.0171	95	0.86	035.43	0
035	Sodium (%)	0904	0.6150	0.0100	0.5808	0.0351	0.0171	95	0.97	035.33	0
035	Sodium (%)	0692	0.6150	0.0100	0.5808	0.0351	0.0171	95	0.97	035.42	0
035	Sodium (%)	0010	0.6150	0.0300	0.5808	0.0351	0.0171	95	0.97	035.43	0
035	Sodium (%)	0510	0.6180	0.0020	0.5808	0.0351	0.0171	95	1.06	035.43	0
035	Sodium (%)	0226	0.6200	0.0000	0.5808	0.0351	0.0171	95	1.12	035.41	0
035	Sodium (%)	0190	0.6200	0.0200	0.5808	0.0351	0.0171	95	1.12	035.42	0
035	Sodium (%)	0066	0.6230	0.0100	0.5808	0.0351	0.0171	95	1.20	035.41	0
035	Sodium (%)	0610	0.6250	0.0100	0.5808	0.0351	0.0171	95	1.26	035.43	0
035	Sodium (%)	0520	0.6281	0.0205	0.5808	0.0351	0.0171	95	1.35	035.41	0
035	Sodium (%)	2124	0.6300	0.0000	0.5808	0.0351	0.0171	95	1.40	035.43	0
035	Sodium (%)	0358	0.6350	0.0300	0.5808	0.0351	0.0171	95	1.54	035.41	0
035	Sodium (%)	0726	0.6400	0.0000	0.5808	0.0351	0.0171	95	1.69	035.42	0
035	Sodium (%)	0045	0.6445	0.0650	0.5808	0.0351	0.0171	95	1.82	035.42	0
035	Sodium (%)	0964	0.6588	0.0831	0.5808	0.0351	0.0171	95	2.22	035.43	0
035	Sodium (%)	0405	0.6735	0.0210	0.5808	0.0351	0.0171	95	2.64	035.41	0
035	Sodium (%)	2234	0.8200	0.0400	0.5808	0.0351	0.0171	95	6.82	035.43	0
035	Sodium (%)	2192	0.5500	0.1000	0.5808	0.0351	0.0171	95	-0.88	035.43	1
035	Sodium (%)	2314	0.6500	0.1000	0.5808	0.0351	0.0171	95	1.97	035.43	1
035	Sodium (%)	2207	0.7100	0.1200	0.5808	0.0351	0.0171	95	3.68	035.42	1
035	Sodium (%)	0620	1.009	0.0108	0.5808	0.0351	0.0171	95	12.20	035.31	2
036	Sulfur (%)	0242	0.3500	0.0000	0.3903	0.0228	0.0120	39	-1.77	036.99	0
036	Sulfur (%)	0226	0.3550	0.0300	0.3903	0.0228	0.0120	39	-1.55	036.04	0
036	Sulfur (%)	0366	0.3550	0.0100	0.3903	0.0228	0.0120	39	-1.55	036.42	0
036	Sulfur (%)	0045	0.3565	0.0230	0.3903	0.0228	0.0120	39	-1.49	036.42	0
036	Sulfur (%)	0297	0.3650	0.0300	0.3903	0.0228	0.0120	39	-1.11	036.43	0
036	Sulfur (%)	0294	0.3700	0.0000	0.3903	0.0228	0.0120	39	-0.89	036.42	0
036	Sulfur (%)	0692	0.3700	0.0400	0.3903	0.0228	0.0120	39	-0.89	036.42	0
036	Sulfur (%)	0870	0.3715	0.0103	0.3903	0.0228	0.0120	39	-0.83	036.42	0
036	Sulfur (%)	0265	0.3750	0.0100	0.3903	0.0228	0.0120	39	-0.67	036.42	0
036	Sulfur (%)	2129	0.3783	0.0165	0.3903	0.0228	0.0120	39	-0.53	036.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
036	Sulfur (%)	0693	0.3790	0.0140	0.3903	0.0228	0.0120	39	-0.50	036.42	0
036	Sulfur (%)	0035	0.3800	0.0200	0.3903	0.0228	0.0120	39	-0.45	036.43	0
036	Sulfur (%)	0083	0.3800	0.0000	0.3903	0.0228	0.0120	39	-0.45	036.43	0
036	Sulfur (%)	0910	0.3805	0.0012	0.3903	0.0228	0.0120	39	-0.43	036.52	0
036	Sulfur (%)	0148	0.3850	0.0100	0.3903	0.0228	0.0120	39	-0.23	036.04	0
036	Sulfur (%)	0035	0.3850	0.0100	0.3903	0.0228	0.0120	39	-0.23	036.42	0
036	Sulfur (%)	0278	0.3850	0.0100	0.3903	0.0228	0.0120	39	-0.23	036.42	0
036	Sulfur (%)	0407	0.3865	0.0148	0.3903	0.0228	0.0120	39	-0.17	036.42	0
036	Sulfur (%)	0357	0.3872	0.0010	0.3903	0.0228	0.0120	39	-0.14	036.42	0
036	Sulfur (%)	0186	0.3919	0.0008	0.3903	0.0228	0.0120	39	0.07	036.42	0
036	Sulfur (%)	0345	0.3925	0.0150	0.3903	0.0228	0.0120	39	0.10	036.43	0
036	Sulfur (%)	0407	0.3930	0.0240	0.3903	0.0228	0.0120	39	0.12	036.43	0
036	Sulfur (%)	0038	0.3938	0.0053	0.3903	0.0228	0.0120	39	0.15	036.00	0
036	Sulfur (%)	0186	0.3939	0.0076	0.3903	0.0228	0.0120	39	0.16	036.52	0
036	Sulfur (%)	0164	0.3950	0.0100	0.3903	0.0228	0.0120	39	0.21	036.42	0
036	Sulfur (%)	0015	0.3970	0.0020	0.3903	0.0228	0.0120	39	0.29	036.53	0
036	Sulfur (%)	2053	0.4000	0.0000	0.3903	0.0228	0.0120	39	0.43	036.01	0
036	Sulfur (%)	0726	0.4000	0.0000	0.3903	0.0228	0.0120	39	0.43	036.42	0
036	Sulfur (%)	0169	0.4000	0.0200	0.3903	0.0228	0.0120	39	0.43	036.43	0
036	Sulfur (%)	0918	0.4005	0.0010	0.3903	0.0228	0.0120	39	0.45	036.43	0
036	Sulfur (%)	0202	0.4025	0.0330	0.3903	0.0228	0.0120	39	0.54	036.43	0
036	Sulfur (%)	0353	0.4100	0.0200	0.3903	0.0228	0.0120	39	0.86	036.43	0
036	Sulfur (%)	0708	0.4135	0.0030	0.3903	0.0228	0.0120	39	1.02	036.42	0
036	Sulfur (%)	0553	0.4140	0.0160	0.3903	0.0228	0.0120	39	1.04	036.53	0
036	Sulfur (%)	0037	0.4240	0.0020	0.3903	0.0228	0.0120	39	1.48	036.42	0
036	Sulfur (%)	0098	0.4287	0.0197	0.3903	0.0228	0.0120	39	1.68	036.43	0
036	Sulfur (%)	0964	0.4317	0.0182	0.3903	0.0228	0.0120	39	1.82	036.43	0
036	Sulfur (%)	2089	0.4350	0.0100	0.3903	0.0228	0.0120	39	1.96	036.43	0
036	Sulfur (%)	0510	0.4550	0.0100	0.3903	0.0228	0.0120	39	2.84	036.43	0
036	Sulfur (%)	0042	0.4075	0.0910	0.3903	0.0228	0.0120	39	0.75	036.42	1
036	Sulfur (%)	0032	3.816	0.0990	0.3903	0.0228	0.0120	39	150.46	036.43	2
037	Zinc (ppm)	2055	138.0	4.000	172.8	14.90	4.779	90	-2.33	037.53	0
037	Zinc (ppm)	0653	142.1	1.000	172.8	14.90	4.779	90	-2.06	037.41	0
037	Zinc (ppm)	0366	147.0	2.000	172.8	14.90	4.779	90	-1.73	037.42	0
037	Zinc (ppm)	0186	148.5	1.000	172.8	14.90	4.779	90	-1.63	037.42	0
037	Zinc (ppm)	0536	148.7	12.95	172.8	14.90	4.779	90	-1.61	037.31	0
037	Zinc (ppm)	0242	149.0	8.000	172.8	14.90	4.779	90	-1.59	037.99	0
037	Zinc (ppm)	2146	150.2	2.080	172.8	14.90	4.779	90	-1.51	037.31	0
037	Zinc (ppm)	0425	153.2	1.240	172.8	14.90	4.779	90	-1.32	037.43	0
037	Zinc (ppm)	2129	153.7	1.100	172.8	14.90	4.779	90	-1.28	037.42	0
037	Zinc (ppm)	0294	155.1	1.300	172.8	14.90	4.779	90	-1.19	037.42	0
037	Zinc (ppm)	0042	155.5	13.00	172.8	14.90	4.779	90	-1.16	037.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
037	Zinc (ppm)	0910	158.0	8.000	172.8	14.90	4.779	90	-0.99	037.41	0
037	Zinc (ppm)	0619	158.2	0.4460	172.8	14.90	4.779	90	-0.98	037.99	0
037	Zinc (ppm)	0918	159.3	2.070	172.8	14.90	4.779	90	-0.90	037.43	0
037	Zinc (ppm)	0017	159.7	13.13	172.8	14.90	4.779	90	-0.88	037.43	0
037	Zinc (ppm)	0169	160.0	10.00	172.8	14.90	4.779	90	-0.86	037.43	0
037	Zinc (ppm)	0164	160.0	10.00	172.8	14.90	4.779	90	-0.86	037.44	0
037	Zinc (ppm)	0226	160.5	2.190	172.8	14.90	4.779	90	-0.82	037.41	0
037	Zinc (ppm)	0009	161.0	2.000	172.8	14.90	4.779	90	-0.79	037.42	0
037	Zinc (ppm)	2012	161.1	0.3900	172.8	14.90	4.779	90	-0.78	037.41	0
037	Zinc (ppm)	0689	161.3	7.000	172.8	14.90	4.779	90	-0.77	037.31	0
037	Zinc (ppm)	0045	162.5	1.000	172.8	14.90	4.779	90	-0.69	037.42	0
037	Zinc (ppm)	0066	162.6	3.457	172.8	14.90	4.779	90	-0.68	037.44	0
037	Zinc (ppm)	0265	163.0	0.0000	172.8	14.90	4.779	90	-0.66	037.42	0
037	Zinc (ppm)	0939	163.0	6.000	172.8	14.90	4.779	90	-0.66	037.53	0
037	Zinc (ppm)	0001	164.0	7.540	172.8	14.90	4.779	90	-0.59	037.31	0
037	Zinc (ppm)	0354	164.6	3.640	172.8	14.90	4.779	90	-0.55	037.41	0
037	Zinc (ppm)	0051	164.8	1.650	172.8	14.90	4.779	90	-0.53	037.41	0
037	Zinc (ppm)	0692	165.0	22.00	172.8	14.90	4.779	90	-0.52	037.42	0
037	Zinc (ppm)	0297	166.5	19.00	172.8	14.90	4.779	90	-0.42	037.43	0
037	Zinc (ppm)	0939	166.7	4.550	172.8	14.90	4.779	90	-0.40	037.31	0
037	Zinc (ppm)	0345	167.0	4.000	172.8	14.90	4.779	90	-0.39	037.43	0
037	Zinc (ppm)	0848	167.9	1.730	172.8	14.90	4.779	90	-0.33	037.41	0
037	Zinc (ppm)	2146	168.1	0.4000	172.8	14.90	4.779	90	-0.31	037.51	0
037	Zinc (ppm)	0263	168.8	1.526	172.8	14.90	4.779	90	-0.27	037.42	0
037	Zinc (ppm)	0693	168.9	5.116	172.8	14.90	4.779	90	-0.26	037.42	0
037	Zinc (ppm)	0968	169.5	1.000	172.8	14.90	4.779	90	-0.22	037.43	0
037	Zinc (ppm)	2114	169.9	0.6400	172.8	14.90	4.779	90	-0.19	037.44	0
037	Zinc (ppm)	0010	170.0	0.0000	172.8	14.90	4.779	90	-0.19	037.33	0
037	Zinc (ppm)	0610	170.0	4.000	172.8	14.90	4.779	90	-0.19	037.43	0
037	Zinc (ppm)	0032	170.2	4.200	172.8	14.90	4.779	90	-0.17	037.43	0
037	Zinc (ppm)	0868	171.0	4.000	172.8	14.90	4.779	90	-0.12	037.31	0
037	Zinc (ppm)	0358	171.4	11.15	172.8	14.90	4.779	90	-0.09	037.41	0
037	Zinc (ppm)	2089	171.5	5.750	172.8	14.90	4.779	90	-0.08	037.43	0
037	Zinc (ppm)	0529	171.7	2.700	172.8	14.90	4.779	90	-0.07	037.31	0
037	Zinc (ppm)	0878	174.0	4.000	172.8	14.90	4.779	90	0.08	037.41	0
037	Zinc (ppm)	0083	174.5	3.000	172.8	14.90	4.779	90	0.12	037.43	0
037	Zinc (ppm)	2192	175.3	3.940	172.8	14.90	4.779	90	0.17	037.43	0
037	Zinc (ppm)	0921	175.5	1.000	172.8	14.90	4.779	90	0.18	037.31	0
037	Zinc (ppm)	0003	175.5	1.000	172.8	14.90	4.779	90	0.18	037.41	0
037	Zinc (ppm)	0884	175.9	11.30	172.8	14.90	4.779	90	0.21	037.31	0
037	Zinc (ppm)	0038	176.3	7.700	172.8	14.90	4.779	90	0.23	037.32	0
037	Zinc (ppm)	0407	176.4	4.222	172.8	14.90	4.779	90	0.24	037.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
037	Zinc (ppm)	0175	176.5	3.000	172.8	14.90	4.779	90	0.25	037.31	0
037	Zinc (ppm)	2188	176.8	1.900	172.8	14.90	4.779	90	0.27	037.31	0
037	Zinc (ppm)	2124	176.8	1.100	172.8	14.90	4.779	90	0.27	037.43	0
037	Zinc (ppm)	0682	176.8	0.0000	172.8	14.90	4.779	90	0.27	037.43	0
037	Zinc (ppm)	0510	177.0	0.0000	172.8	14.90	4.779	90	0.28	037.43	0
037	Zinc (ppm)	2196	177.3	0.0000	172.8	14.90	4.779	90	0.31	037.31	0
037	Zinc (ppm)	0202	177.7	16.80	172.8	14.90	4.779	90	0.33	037.43	0
037	Zinc (ppm)	0357	178.2	0.2910	172.8	14.90	4.779	90	0.36	037.42	0
037	Zinc (ppm)	0186	178.5	3.000	172.8	14.90	4.779	90	0.39	037.52	0
037	Zinc (ppm)	0590	178.5	1.000	172.8	14.90	4.779	90	0.39	037.99	0
037	Zinc (ppm)	0098	179.0	0.5000	172.8	14.90	4.779	90	0.42	037.41	0
037	Zinc (ppm)	0019	179.5	5.100	172.8	14.90	4.779	90	0.45	037.41	0
037	Zinc (ppm)	0520	179.5	13.00	172.8	14.90	4.779	90	0.45	037.41	0
037	Zinc (ppm)	0948	180.0	0.9380	172.8	14.90	4.779	90	0.48	037.33	0
037	Zinc (ppm)	0148	180.0	4.000	172.8	14.90	4.779	90	0.49	037.41	0
037	Zinc (ppm)	2115	180.4	0.1400	172.8	14.90	4.779	90	0.51	037.31	0
037	Zinc (ppm)	0015	181.2	24.64	172.8	14.90	4.779	90	0.57	037.53	0
037	Zinc (ppm)	0353	181.5	1.000	172.8	14.90	4.779	90	0.59	037.43	0
037	Zinc (ppm)	0553	182.0	4.000	172.8	14.90	4.779	90	0.62	037.53	0
037	Zinc (ppm)	0190	182.7	1.300	172.8	14.90	4.779	90	0.66	037.42	0
037	Zinc (ppm)	0407	183.9	0.4380	172.8	14.90	4.779	90	0.75	037.43	0
037	Zinc (ppm)	0726	185.1	0.3000	172.8	14.90	4.779	90	0.82	037.42	0
037	Zinc (ppm)	2022	188.4	6.303	172.8	14.90	4.779	90	1.05	037.31	0
037	Zinc (ppm)	0964	189.5	24.86	172.8	14.90	4.779	90	1.12	037.43	0
037	Zinc (ppm)	0035	193.0	2.000	172.8	14.90	4.779	90	1.36	037.42	0
037	Zinc (ppm)	0100	193.0	0.0000	172.8	14.90	4.779	90	1.36	037.99	0
037	Zinc (ppm)	0035	193.5	1.000	172.8	14.90	4.779	90	1.39	037.43	0
037	Zinc (ppm)	0598	195.3	1.800	172.8	14.90	4.779	90	1.51	037.41	0
037	Zinc (ppm)	0278	195.8	4.400	172.8	14.90	4.779	90	1.55	037.42	0
037	Zinc (ppm)	0098	197.3	13.00	172.8	14.90	4.779	90	1.65	037.43	0
037	Zinc (ppm)	0037	198.4	1.500	172.8	14.90	4.779	90	1.72	037.42	0
037	Zinc (ppm)	2234	201.8	3.250	172.8	14.90	4.779	90	1.95	037.43	0
037	Zinc (ppm)	0033	202.0	2.000	172.8	14.90	4.779	90	1.96	037.43	0
037	Zinc (ppm)	0405	206.5	1.400	172.8	14.90	4.779	90	2.26	037.41	0
037	Zinc (ppm)	0870	226.0	2.300	172.8	14.90	4.779	90	3.57	037.42	0
037	Zinc (ppm)	0620	279.7	24.00	172.8	14.90	4.779	90	7.18	037.31	0
037	Zinc (ppm)	2314	336.5	5.700	172.8	14.90	4.779	90	10.99	037.43	0
037	Zinc (ppm)	0572	165.5	27.00	172.8	14.90	4.779	90	-0.49	037.53	1
037	Zinc (ppm)	2207	190.0	30.00	172.8	14.90	4.779	90	1.16	037.42	1
037	Zinc (ppm)	0914	0.1700	0.0260	172.8	14.90	4.779	90	-11.58	037.53	2
038	Molybdenum (ppm)	0353	0.7048	0.0582	0.9898	0.2073	0.0575	19	-1.37	038.43	0
038	Molybdenum (ppm)	0510	0.8000	0.0000	0.9898	0.2073	0.0575	19	-0.92	038.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
038	Molybdenum (ppm)	0407	0.8088	0.0029	0.9898	0.2073	0.0575	19	-0.87	038.41	0
038	Molybdenum (ppm)	2055	0.8150	0.0300	0.9898	0.2073	0.0575	19	-0.84	038.53	0
038	Molybdenum (ppm)	0169	0.8200	0.0200	0.9898	0.2073	0.0575	19	-0.82	038.43	0
038	Molybdenum (ppm)	0407	0.8420	0.0160	0.9898	0.2073	0.0575	19	-0.71	038.43	0
038	Molybdenum (ppm)	0345	0.8550	0.0300	0.9898	0.2073	0.0575	19	-0.65	038.43	0
038	Molybdenum (ppm)	0964	0.9000	0.0290	0.9898	0.2073	0.0575	19	-0.43	038.43	0
038	Molybdenum (ppm)	0278	0.9150	0.0500	0.9898	0.2073	0.0575	19	-0.36	038.42	0
038	Molybdenum (ppm)	0910	0.9700	0.0800	0.9898	0.2073	0.0575	19	-0.10	038.52	0
038	Molybdenum (ppm)	0553	1.002	0.0370	0.9898	0.2073	0.0575	19	0.06	038.53	0
038	Molybdenum (ppm)	0164	1.040	0.0200	0.9898	0.2073	0.0575	19	0.24	038.53	0
038	Molybdenum (ppm)	0918	1.058	0.0094	0.9898	0.2073	0.0575	19	0.33	038.53	0
038	Molybdenum (ppm)	0226	1.085	0.0300	0.9898	0.2073	0.0575	19	0.46	038.41	0
038	Molybdenum (ppm)	0297	1.195	0.3900	0.9898	0.2073	0.0575	19	0.99	038.43	0
038	Molybdenum (ppm)	0015	1.195	0.1900	0.9898	0.2073	0.0575	19	0.99	038.53	0
038	Molybdenum (ppm)	0083	1.200	0.0000	0.9898	0.2073	0.0575	19	1.01	038.43	0
038	Molybdenum (ppm)	0265	1.370	0.0800	0.9898	0.2073	0.0575	19	1.83	038.42	0
038	Molybdenum (ppm)	0045	1.440	0.0200	0.9898	0.2073	0.0575	19	2.17	038.42	0
038	Molybdenum (ppm)	0693	0.9960	0.4780	0.9898	0.2073	0.0575	19	0.03	038.42	1
038	Molybdenum (ppm)	0202	< 2		0.9898	0.2073	0.0575	19		038.43	5
040	Barium (ppm)	0015	6.044	0.7070				2		040.53	0
040	Barium (ppm)	0918	6.258	0.0120				2		040.53	0
041	Vanadium (ppm)	0553	0.1995	0.0010				1		041.53	0
042	Chloride (%)	0948	0.7290	0.0000				3		042.00	0
042	Chloride (%)	2124	0.7295	0.0010				3		042.00	0
042	Chloride (%)	0297	0.7850	0.0500				3		042.99	0
101	Choline Chloride (ppm)	0227	3,980	20.00				1		101.99	0
102	Niacin (ppm)	0227	133.5	5.000				2		102.01	0
102	Niacin (ppm)	2124	143.2	0.0100				2		102.02	0
103	Pantothenic Acid (ppm)	0227	38.15	0.5000				2		103.01	0
103	Pantothenic Acid (ppm)	2124	94.90	2.800				2		103.02	0
104	Riboflavin (ppm)	0910	13.99	2.400				3		104.03	0
104	Riboflavin (ppm)	2124	17.93	0.0700				3		104.03	0
104	Riboflavin (ppm)	0227	20.65	0.1000				3		104.00	0
105	Thiamine (ppm)	0227	188.5	7.000				3		105.01	0
105	Thiamine (ppm)	0910	215.7	28.89				3		105.00	0
105	Thiamine (ppm)	2124	232.6	0.9700				3		105.00	0
106	Vitamin A (KU / kg)	0905	8.581	1.606	13.05	2.790	0.8493	22	-1.60	106.02	0
106	Vitamin A (KU / kg)	0860	8.810	0.0000	13.05	2.790	0.8493	22	-1.52	106.02	0
106	Vitamin A (KU / kg)	0904	9.610	0.6800	13.05	2.790	0.8493	22	-1.23	106.99	0
106	Vitamin A (KU / kg)	0035	9.700	0.1600	13.05	2.790	0.8493	22	-1.20	106.02	0
106	Vitamin A (KU / kg)	0038	10.83	0.1096	13.05	2.790	0.8493	22	-0.80	106.02	0
106	Vitamin A (KU / kg)	0914	11.27	0.1200	13.05	2.790	0.8493	22	-0.64	106.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
106	Vitamin A (KU / kg)	0148	11.53	0.6560	13.05	2.790	0.8493	22	-0.55	106.02	0
106	Vitamin A (KU / kg)	2053	11.95	1.900	13.05	2.790	0.8493	22	-0.39	106.02	0
106	Vitamin A (KU / kg)	0010	12.91	0.3300	13.05	2.790	0.8493	22	-0.05	106.02	0
106	Vitamin A (KU / kg)	0941	13.07	0.8840	13.05	2.790	0.8493	22	0.01	106.02	0
106	Vitamin A (KU / kg)	2022	13.50	1.000	13.05	2.790	0.8493	22	0.16	106.02	0
106	Vitamin A (KU / kg)	0610	13.60	1.000	13.05	2.790	0.8493	22	0.20	106.02	0
106	Vitamin A (KU / kg)	0676	13.88	0.1500	13.05	2.790	0.8493	22	0.30	106.02	0
106	Vitamin A (KU / kg)	0910	14.20	0.8000	13.05	2.790	0.8493	22	0.41	106.02	0
106	Vitamin A (KU / kg)	0512	14.28	0.1000	13.05	2.790	0.8493	22	0.44	106.02	0
106	Vitamin A (KU / kg)	0169	14.35	0.1000	13.05	2.790	0.8493	22	0.47	106.02	0
106	Vitamin A (KU / kg)	0227	14.60	0.2000	13.05	2.790	0.8493	22	0.56	106.00	0
106	Vitamin A (KU / kg)	0689	14.90	0.2000	13.05	2.790	0.8493	22	0.66	106.02	0
106	Vitamin A (KU / kg)	0036	15.00	0.8000	13.05	2.790	0.8493	22	0.70	106.02	0
106	Vitamin A (KU / kg)	0098	15.70	1.200	13.05	2.790	0.8493	22	0.95	106.01	0
106	Vitamin A (KU / kg)	0675	18.16	4.190	13.05	2.790	0.8493	22	1.83	106.02	0
106	Vitamin A (KU / kg)	0619	20.05	2.500	13.05	2.790	0.8493	22	2.51	106.02	0
106	Vitamin A (KU / kg)	0017	18.63	12.77	13.05	2.790	0.8493	22	2.00	106.02	1
107	Vitamin B12 (ppb)	0227	141.0	12.00				1		107.00	0
108	Vitamin D3 (KU / kg)	0610	0.4500	0.1000	1.988	0.7810	0.3170	10	-1.97	108.02	0
108	Vitamin D3 (KU / kg)	0160	1.305	0.1700	1.988	0.7810	0.3170	10	-0.87	108.99	0
108	Vitamin D3 (KU / kg)	0227	1.690	0.0000	1.988	0.7810	0.3170	10	-0.38	108.02	0
108	Vitamin D3 (KU / kg)	0010	1.695	0.3100	1.988	0.7810	0.3170	10	-0.38	108.99	0
108	Vitamin D3 (KU / kg)	0051	1.715	0.0100	1.988	0.7810	0.3170	10	-0.35	108.02	0
108	Vitamin D3 (KU / kg)	0914	2.040	0.0000	1.988	0.7810	0.3170	10	0.07	108.99	0
108	Vitamin D3 (KU / kg)	0910	2.305	1.070	1.988	0.7810	0.3170	10	0.41	108.02	0
108	Vitamin D3 (KU / kg)	0675	2.500	0.2600	1.988	0.7810	0.3170	10	0.66	108.02	0
108	Vitamin D3 (KU / kg)	0619	2.655	0.2500	1.988	0.7810	0.3170	10	0.85	108.99	0
108	Vitamin D3 (KU / kg)	0169	6.500	1.000	1.988	0.7810	0.3170	10	5.78	108.01	0
109	Vitamin E (IU / kg)	0169	158.5	1.000	317.4	68.56	25.40	17	-2.32	109.02	0
109	Vitamin E (IU / kg)	0038	215.2	9.900	317.4	68.56	25.40	17	-1.49	109.02	0
109	Vitamin E (IU / kg)	0590	261.0	2.000	317.4	68.56	25.40	17	-0.82	109.02	0
109	Vitamin E (IU / kg)	0905	263.0	52.00	317.4	68.56	25.40	17	-0.79	109.02	0
109	Vitamin E (IU / kg)	0675	272.9	20.77	317.4	68.56	25.40	17	-0.65	109.02	0
109	Vitamin E (IU / kg)	2022	300.0	0.0000	317.4	68.56	25.40	17	-0.25	109.02	0
109	Vitamin E (IU / kg)	0036	300.5	6.400	317.4	68.56	25.40	17	-0.25	109.02	0
109	Vitamin E (IU / kg)	0098	306.0	40.00	317.4	68.56	25.40	17	-0.17	109.02	0
109	Vitamin E (IU / kg)	2152	317.5	5.000	317.4	68.56	25.40	17	0.00	109.02	0
109	Vitamin E (IU / kg)	0610	324.5	13.00	317.4	68.56	25.40	17	0.10	109.02	0
109	Vitamin E (IU / kg)	0910	332.5	21.00	317.4	68.56	25.40	17	0.22	109.02	0
109	Vitamin E (IU / kg)	0619	350.5	75.00	317.4	68.56	25.40	17	0.48	109.02	0
109	Vitamin E (IU / kg)	0941	357.9	41.05	317.4	68.56	25.40	17	0.59	109.02	0
109	Vitamin E (IU / kg)	0148	365.6	29.77	317.4	68.56	25.40	17	0.70	109.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
109	Vitamin E (IU / kg)	0676	393.0	1.950	317.4	68.56	25.40	17	1.10	109.02	0
109	Vitamin E (IU / kg)	0227	401.0	2.000	317.4	68.56	25.40	17	1.22	109.02	0
109	Vitamin E (IU / kg)	0160	547.5	111.0	317.4	68.56	25.40	17	3.36	109.99	0
111	Vitamin C, Phosphorylated (ppm)	0148	15.50	0.4000				3		111.01	0
111	Vitamin C, Phosphorylated (ppm)	2314	33.75	0.1000				3		111.01	0
111	Vitamin C, Phosphorylated (ppm)	0227	60.25	0.1000				3		111.00	0
111	Vitamin C, Phosphorylated (ppm)	0676	4.830	1.020				3		111.01	1
111	Vitamin C, Phosphorylated (ppm)	0914	< 10					3		111.01	5
111	Vitamin C, Phosphorylated (ppm)	0160	< 60					3		111.99	5
112	Pyridoxine (µg / g)	0227	12.95	1.100				2		112.01	0
112	Pyridoxine (µg / g)	2124	17.82	1.300				2		112.01	0
113	Folic Acid (ppm)	0227	4.390	0.4000				1		113.01	0
114	Biotin (ppm)	0227	0.3600	0.0360				1		114.01	0
118	Peroxide value (meq/kg)	2081	2.680	1.600				1		118.99	0
120	Alanine (%)	0889	1.485	0.0100	1.975	0.0854	0.0367	30	-5.74	120.99	0
120	Alanine (%)	0353	1.720	0.0200	1.975	0.0854	0.0367	30	-2.99	120.00	0
120	Alanine (%)	0676	1.765	0.1300	1.975	0.0854	0.0367	30	-2.46	120.05	0
120	Alanine (%)	0675	1.850	0.0200	1.975	0.0854	0.0367	30	-1.47	120.00	0
120	Alanine (%)	0619	1.885	0.0300	1.975	0.0854	0.0367	30	-1.06	120.00	0
120	Alanine (%)	0918	1.895	0.0197	1.975	0.0854	0.0367	30	-0.94	120.00	0
120	Alanine (%)	0148	1.915	0.0700	1.975	0.0854	0.0367	30	-0.71	120.05	0
120	Alanine (%)	0682	1.939	0.0000	1.975	0.0854	0.0367	30	-0.43	120.00	0
120	Alanine (%)	2115	1.950	0.0060	1.975	0.0854	0.0367	30	-0.30	120.99	0
120	Alanine (%)	0610	1.955	0.0100	1.975	0.0854	0.0367	30	-0.24	120.05	0
120	Alanine (%)	0098	1.966	0.0120	1.975	0.0854	0.0367	30	-0.11	120.02	0
120	Alanine (%)	2059	1.970	0.0050	1.975	0.0854	0.0367	30	-0.07	120.00	0
120	Alanine (%)	0652	1.970	0.0600	1.975	0.0854	0.0367	30	-0.06	120.00	0
120	Alanine (%)	0571	1.972	0.0130	1.975	0.0854	0.0367	30	-0.05	120.00	0
120	Alanine (%)	0941	1.974	0.1030	1.975	0.0854	0.0367	30	-0.02	120.00	0
120	Alanine (%)	0859	1.975	0.0126	1.975	0.0854	0.0367	30	-0.01	120.00	0
120	Alanine (%)	0870	1.978	0.0646	1.975	0.0854	0.0367	30	0.03	120.00	0
120	Alanine (%)	0644	1.987	0.0030	1.975	0.0854	0.0367	30	0.13	120.00	0
120	Alanine (%)	0910	1.995	0.0300	1.975	0.0854	0.0367	30	0.23	120.00	0
120	Alanine (%)	2188	2.002	0.0040	1.975	0.0854	0.0367	30	0.31	120.05	0
120	Alanine (%)	0872	2.004	0.0090	1.975	0.0854	0.0367	30	0.33	120.00	0
120	Alanine (%)	0968	2.005	0.0300	1.975	0.0854	0.0367	30	0.35	120.00	0
120	Alanine (%)	0878	2.007	0.0100	1.975	0.0854	0.0367	30	0.37	120.00	0
120	Alanine (%)	0939	2.035	0.0100	1.975	0.0854	0.0367	30	0.70	120.00	0
120	Alanine (%)	0626	2.045	0.0610	1.975	0.0854	0.0367	30	0.81	120.05	0
120	Alanine (%)	0227	2.060	0.0400	1.975	0.0854	0.0367	30	0.99	120.99	0
120	Alanine (%)	0868	2.083	0.1080	1.975	0.0854	0.0367	30	1.26	120.00	0
120	Alanine (%)	0407	2.098	0.1180	1.975	0.0854	0.0367	30	1.44	120.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
120	Alanine (%)	0354	2.111	0.0130	1.975	0.0854	0.0367	30	1.58	120.00	0
120	Alanine (%)	0226	2.120	0.0800	1.975	0.0854	0.0367	30	1.69	120.00	0
121	Arginine (%)	0889	1.325	0.0100	1.817	0.0899	0.0244	29	-5.48	121.99	0
121	Arginine (%)	0676	1.630	0.0000	1.817	0.0899	0.0244	29	-2.09	121.05	0
121	Arginine (%)	0918	1.664	0.0003	1.817	0.0899	0.0244	29	-1.71	121.00	0
121	Arginine (%)	0626	1.695	0.1000	1.817	0.0899	0.0244	29	-1.36	121.05	0
121	Arginine (%)	0675	1.705	0.0300	1.817	0.0899	0.0244	29	-1.25	121.00	0
121	Arginine (%)	0619	1.745	0.0100	1.817	0.0899	0.0244	29	-0.81	121.00	0
121	Arginine (%)	0682	1.761	0.0000	1.817	0.0899	0.0244	29	-0.63	121.00	0
121	Arginine (%)	0859	1.765	0.0236	1.817	0.0899	0.0244	29	-0.59	121.00	0
121	Arginine (%)	0870	1.781	0.0193	1.817	0.0899	0.0244	29	-0.40	121.00	0
121	Arginine (%)	0868	1.787	0.0350	1.817	0.0899	0.0244	29	-0.34	121.00	0
121	Arginine (%)	0652	1.795	0.0500	1.817	0.0899	0.0244	29	-0.25	121.00	0
121	Arginine (%)	0407	1.815	0.0360	1.817	0.0899	0.0244	29	-0.03	121.05	0
121	Arginine (%)	0098	1.816	0.0020	1.817	0.0899	0.0244	29	-0.02	121.02	0
121	Arginine (%)	0227	1.820	0.0400	1.817	0.0899	0.0244	29	0.03	121.99	0
121	Arginine (%)	0968	1.825	0.0100	1.817	0.0899	0.0244	29	0.08	121.00	0
121	Arginine (%)	0939	1.830	0.0200	1.817	0.0899	0.0244	29	0.14	121.00	0
121	Arginine (%)	0941	1.831	0.0920	1.817	0.0899	0.0244	29	0.15	121.00	0
121	Arginine (%)	0571	1.851	0.0250	1.817	0.0899	0.0244	29	0.37	121.00	0
121	Arginine (%)	0878	1.852	0.0040	1.817	0.0899	0.0244	29	0.38	121.00	0
121	Arginine (%)	0644	1.859	0.0030	1.817	0.0899	0.0244	29	0.46	121.00	0
121	Arginine (%)	0872	1.867	0.0070	1.817	0.0899	0.0244	29	0.55	121.00	0
121	Arginine (%)	2059	1.868	0.0030	1.817	0.0899	0.0244	29	0.56	121.00	0
121	Arginine (%)	0910	1.870	0.0400	1.817	0.0899	0.0244	29	0.58	121.00	0
121	Arginine (%)	0353	1.880	0.0000	1.817	0.0899	0.0244	29	0.70	121.00	0
121	Arginine (%)	0354	1.904	0.0010	1.817	0.0899	0.0244	29	0.96	121.00	0
121	Arginine (%)	2115	1.915	0.0120	1.817	0.0899	0.0244	29	1.09	121.99	0
121	Arginine (%)	0610	1.920	0.0200	1.817	0.0899	0.0244	29	1.14	121.05	0
121	Arginine (%)	0226	2.035	0.1100	1.817	0.0899	0.0244	29	2.42	121.00	0
121	Arginine (%)	2188	2.147	0.0030	1.817	0.0899	0.0244	29	3.66	121.05	0
121	Arginine (%)	0148	1.755	0.1900	1.817	0.0899	0.0244	29	-0.69	121.05	1
122	Aspartic (%)	0676	2.260	0.1400	2.593	0.1045	0.0465	30	-3.19	122.05	0
122	Aspartic (%)	0353	2.295	0.0900	2.593	0.1045	0.0465	30	-2.85	122.00	0
122	Aspartic (%)	0626	2.405	0.0100	2.593	0.1045	0.0465	30	-1.80	122.05	0
122	Aspartic (%)	0675	2.445	0.1100	2.593	0.1045	0.0465	30	-1.42	122.00	0
122	Aspartic (%)	0918	2.473	0.0446	2.593	0.1045	0.0465	30	-1.15	122.00	0
122	Aspartic (%)	0941	2.505	0.0910	2.593	0.1045	0.0465	30	-0.85	122.00	0
122	Aspartic (%)	0859	2.509	0.0243	2.593	0.1045	0.0465	30	-0.81	122.00	0
122	Aspartic (%)	0407	2.523	0.1150	2.593	0.1045	0.0465	30	-0.68	122.05	0
122	Aspartic (%)	0682	2.535	0.0000	2.593	0.1045	0.0465	30	-0.56	122.00	0
122	Aspartic (%)	2115	2.558	0.0760	2.593	0.1045	0.0465	30	-0.34	122.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
122	Aspartic (%)	0968	2.565	0.0300	2.593	0.1045	0.0465	30	-0.27	122.00	0
122	Aspartic (%)	2059	2.572	0.0070	2.593	0.1045	0.0465	30	-0.21	122.00	0
122	Aspartic (%)	0571	2.591	0.0480	2.593	0.1045	0.0465	30	-0.02	122.00	0
122	Aspartic (%)	0619	2.605	0.0100	2.593	0.1045	0.0465	30	0.11	122.00	0
122	Aspartic (%)	0098	2.608	0.0060	2.593	0.1045	0.0465	30	0.14	122.02	0
122	Aspartic (%)	0644	2.613	0.0030	2.593	0.1045	0.0465	30	0.18	122.00	0
122	Aspartic (%)	0872	2.616	0.0300	2.593	0.1045	0.0465	30	0.22	122.00	0
122	Aspartic (%)	0870	2.622	0.0541	2.593	0.1045	0.0465	30	0.27	122.00	0
122	Aspartic (%)	0652	2.630	0.1200	2.593	0.1045	0.0465	30	0.35	122.00	0
122	Aspartic (%)	0227	2.640	0.0400	2.593	0.1045	0.0465	30	0.45	122.99	0
122	Aspartic (%)	0878	2.641	0.0020	2.593	0.1045	0.0465	30	0.46	122.00	0
122	Aspartic (%)	2188	2.646	0.0040	2.593	0.1045	0.0465	30	0.50	122.05	0
122	Aspartic (%)	0939	2.660	0.0200	2.593	0.1045	0.0465	30	0.64	122.00	0
122	Aspartic (%)	0610	2.670	0.0400	2.593	0.1045	0.0465	30	0.73	122.05	0
122	Aspartic (%)	0910	2.675	0.1100	2.593	0.1045	0.0465	30	0.78	122.00	0
122	Aspartic (%)	0148	2.680	0.0200	2.593	0.1045	0.0465	30	0.83	122.05	0
122	Aspartic (%)	0354	2.681	0.0140	2.593	0.1045	0.0465	30	0.84	122.00	0
122	Aspartic (%)	0868	2.729	0.0550	2.593	0.1045	0.0465	30	1.29	122.00	0
122	Aspartic (%)	0889	2.805	0.0100	2.593	0.1045	0.0465	30	2.03	122.99	0
122	Aspartic (%)	0226	2.885	0.0700	2.593	0.1045	0.0465	30	2.79	122.00	0
124	Cysteine/Cystine (%)	0675	0.3400	0.0400	0.4496	0.0332	0.0085	29	-3.30	124.00	0
124	Cysteine/Cystine (%)	0870	0.3752	0.0008	0.4496	0.0332	0.0085	29	-2.24	124.00	0
124	Cysteine/Cystine (%)	0889	0.4125	0.0050	0.4496	0.0332	0.0085	29	-1.12	124.99	0
124	Cysteine/Cystine (%)	0910	0.4200	0.0200	0.4496	0.0332	0.0085	29	-0.89	124.00	0
124	Cysteine/Cystine (%)	0227	0.4200	0.0000	0.4496	0.0332	0.0085	29	-0.89	124.99	0
124	Cysteine/Cystine (%)	0859	0.4205	0.0068	0.4496	0.0332	0.0085	29	-0.88	124.00	0
124	Cysteine/Cystine (%)	0941	0.4255	0.0070	0.4496	0.0332	0.0085	29	-0.73	124.00	0
124	Cysteine/Cystine (%)	0644	0.4285	0.0050	0.4496	0.0332	0.0085	29	-0.64	124.00	0
124	Cysteine/Cystine (%)	0918	0.4300	0.0200	0.4496	0.0332	0.0085	29	-0.59	124.00	0
124	Cysteine/Cystine (%)	0878	0.4320	0.0040	0.4496	0.0332	0.0085	29	-0.53	124.00	0
124	Cysteine/Cystine (%)	0652	0.4400	0.0000	0.4496	0.0332	0.0085	29	-0.29	124.00	0
124	Cysteine/Cystine (%)	2059	0.4400	0.0040	0.4496	0.0332	0.0085	29	-0.29	124.00	0
124	Cysteine/Cystine (%)	0939	0.4450	0.0300	0.4496	0.0332	0.0085	29	-0.14	124.00	0
124	Cysteine/Cystine (%)	0148	0.4490	0.0180	0.4496	0.0332	0.0085	29	-0.02	124.05	0
124	Cysteine/Cystine (%)	0868	0.4510	0.0080	0.4496	0.0332	0.0085	29	0.04	124.00	0
124	Cysteine/Cystine (%)	2115	0.4510	0.0060	0.4496	0.0332	0.0085	29	0.04	124.99	0
124	Cysteine/Cystine (%)	0872	0.4515	0.0090	0.4496	0.0332	0.0085	29	0.06	124.00	0
124	Cysteine/Cystine (%)	0571	0.4535	0.0070	0.4496	0.0332	0.0085	29	0.12	124.00	0
124	Cysteine/Cystine (%)	0676	0.4565	0.0010	0.4496	0.0332	0.0085	29	0.21	124.05	0
124	Cysteine/Cystine (%)	2188	0.4570	0.0040	0.4496	0.0332	0.0085	29	0.22	124.05	0
124	Cysteine/Cystine (%)	0968	0.4580	0.0020	0.4496	0.0332	0.0085	29	0.25	124.00	0
124	Cysteine/Cystine (%)	0354	0.4650	0.0040	0.4496	0.0332	0.0085	29	0.46	124.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
124	Cysteine/Cystine (%)	0610	0.4700	0.0000	0.4496	0.0332	0.0085	29	0.61	124.05	0
124	Cysteine/Cystine (%)	0098	0.4715	0.0110	0.4496	0.0332	0.0085	29	0.66	124.02	0
124	Cysteine/Cystine (%)	0682	0.4940	0.0000	0.4496	0.0332	0.0085	29	1.34	124.00	0
124	Cysteine/Cystine (%)	0619	0.4985	0.0050	0.4496	0.0332	0.0085	29	1.47	124.00	0
124	Cysteine/Cystine (%)	0226	0.5250	0.0300	0.4496	0.0332	0.0085	29	2.27	124.00	0
124	Cysteine/Cystine (%)	0353	0.5500	0.0000	0.4496	0.0332	0.0085	29	3.03	124.00	0
124	Cysteine/Cystine (%)	0848	0.6800	0.0000	0.4496	0.0332	0.0085	29	6.94	124.00	0
124	Cysteine/Cystine (%)	0407	0.8990	0.0720	0.4496	0.0332	0.0085	29	13.54	124.05	1
125	Glutamic (%)	2115	4.695	0.0710	5.305	0.2301	0.0872	29	-2.65	125.99	0
125	Glutamic (%)	0675	4.880	0.0800	5.305	0.2301	0.0872	29	-1.85	125.00	0
125	Glutamic (%)	0918	4.955	0.0123	5.305	0.2301	0.0872	29	-1.52	125.00	0
125	Glutamic (%)	2188	5.052	0.0040	5.305	0.2301	0.0872	29	-1.10	125.05	0
125	Glutamic (%)	0353	5.130	0.1000	5.305	0.2301	0.0872	29	-0.76	125.00	0
125	Glutamic (%)	0676	5.140	0.2600	5.305	0.2301	0.0872	29	-0.72	125.05	0
125	Glutamic (%)	0859	5.181	0.0666	5.305	0.2301	0.0872	29	-0.54	125.00	0
125	Glutamic (%)	2059	5.189	0.0510	5.305	0.2301	0.0872	29	-0.51	125.00	0
125	Glutamic (%)	0968	5.190	0.0200	5.305	0.2301	0.0872	29	-0.50	125.00	0
125	Glutamic (%)	0571	5.211	0.0980	5.305	0.2301	0.0872	29	-0.41	125.00	0
125	Glutamic (%)	0098	5.231	0.0010	5.305	0.2301	0.0872	29	-0.33	125.02	0
125	Glutamic (%)	0941	5.262	0.2770	5.305	0.2301	0.0872	29	-0.19	125.00	0
125	Glutamic (%)	0610	5.265	0.0900	5.305	0.2301	0.0872	29	-0.18	125.05	0
125	Glutamic (%)	0652	5.275	0.0700	5.305	0.2301	0.0872	29	-0.13	125.00	0
125	Glutamic (%)	0354	5.315	0.0030	5.305	0.2301	0.0872	29	0.04	125.00	0
125	Glutamic (%)	0868	5.315	0.0430	5.305	0.2301	0.0872	29	0.04	125.00	0
125	Glutamic (%)	0619	5.320	0.0400	5.305	0.2301	0.0872	29	0.06	125.00	0
125	Glutamic (%)	0644	5.341	0.0220	5.305	0.2301	0.0872	29	0.15	125.00	0
125	Glutamic (%)	0682	5.354	0.0000	5.305	0.2301	0.0872	29	0.21	125.00	0
125	Glutamic (%)	0872	5.366	0.0320	5.305	0.2301	0.0872	29	0.26	125.00	0
125	Glutamic (%)	0870	5.386	0.0731	5.305	0.2301	0.0872	29	0.35	125.00	0
125	Glutamic (%)	0910	5.405	0.1700	5.305	0.2301	0.0872	29	0.43	125.00	0
125	Glutamic (%)	0626	5.423	0.1630	5.305	0.2301	0.0872	29	0.51	125.05	0
125	Glutamic (%)	0227	5.430	0.0800	5.305	0.2301	0.0872	29	0.54	125.99	0
125	Glutamic (%)	0878	5.606	0.0220	5.305	0.2301	0.0872	29	1.31	125.00	0
125	Glutamic (%)	0148	5.640	0.0200	5.305	0.2301	0.0872	29	1.45	125.05	0
125	Glutamic (%)	0226	5.855	0.2700	5.305	0.2301	0.0872	29	2.39	125.00	0
125	Glutamic (%)	0407	5.914	0.3610	5.305	0.2301	0.0872	29	2.64	125.05	0
125	Glutamic (%)	0939	5.965	0.0300	5.305	0.2301	0.0872	29	2.87	125.00	0
126	Glycine (%)	0676	1.740	0.0800	1.945	0.0842	0.0384	29	-2.44	126.05	0
126	Glycine (%)	0353	1.775	0.0500	1.945	0.0842	0.0384	29	-2.02	126.00	0
126	Glycine (%)	0148	1.820	0.1600	1.945	0.0842	0.0384	29	-1.49	126.05	0
126	Glycine (%)	0675	1.845	0.0300	1.945	0.0842	0.0384	29	-1.19	126.00	0
126	Glycine (%)	0682	1.857	0.0000	1.945	0.0842	0.0384	29	-1.05	126.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
126	Glycine (%)	0918	1.860	0.0192	1.945	0.0842	0.0384	29	-1.02	126.00	0
126	Glycine (%)	2115	1.869	0.0060	1.945	0.0842	0.0384	29	-0.91	126.99	0
126	Glycine (%)	0626	1.892	0.0260	1.945	0.0842	0.0384	29	-0.63	126.05	0
126	Glycine (%)	0619	1.900	0.0200	1.945	0.0842	0.0384	29	-0.54	126.00	0
126	Glycine (%)	0859	1.901	0.0076	1.945	0.0842	0.0384	29	-0.53	126.00	0
126	Glycine (%)	0652	1.940	0.0000	1.945	0.0842	0.0384	29	-0.06	126.00	0
126	Glycine (%)	0571	1.942	0.0260	1.945	0.0842	0.0384	29	-0.04	126.00	0
126	Glycine (%)	0868	1.945	0.0180	1.945	0.0842	0.0384	29	0.00	126.00	0
126	Glycine (%)	0878	1.949	0.0100	1.945	0.0842	0.0384	29	0.04	126.00	0
126	Glycine (%)	0941	1.960	0.1120	1.945	0.0842	0.0384	29	0.17	126.00	0
126	Glycine (%)	0968	1.965	0.0100	1.945	0.0842	0.0384	29	0.23	126.00	0
126	Glycine (%)	0872	1.973	0.0060	1.945	0.0842	0.0384	29	0.33	126.00	0
126	Glycine (%)	2059	1.974	0.0150	1.945	0.0842	0.0384	29	0.33	126.00	0
126	Glycine (%)	0098	1.981	0.0610	1.945	0.0842	0.0384	29	0.42	126.02	0
126	Glycine (%)	0644	1.981	0.0100	1.945	0.0842	0.0384	29	0.42	126.00	0
126	Glycine (%)	2188	1.983	0.0050	1.945	0.0842	0.0384	29	0.44	126.05	0
126	Glycine (%)	0227	2.005	0.0100	1.945	0.0842	0.0384	29	0.71	126.99	0
126	Glycine (%)	0939	2.010	0.0000	1.945	0.0842	0.0384	29	0.77	126.00	0
126	Glycine (%)	0870	2.012	0.1251	1.945	0.0842	0.0384	29	0.79	126.00	0
126	Glycine (%)	0910	2.015	0.0500	1.945	0.0842	0.0384	29	0.83	126.00	0
126	Glycine (%)	0610	2.020	0.0400	1.945	0.0842	0.0384	29	0.89	126.05	0
126	Glycine (%)	0354	2.036	0.0040	1.945	0.0842	0.0384	29	1.08	126.00	0
126	Glycine (%)	0226	2.145	0.0700	1.945	0.0842	0.0384	29	2.37	126.00	0
126	Glycine (%)	0407	2.410	0.1440	1.945	0.0842	0.0384	29	5.52	126.05	0
126	Glycine (%)	0889	0.6075	0.0150	1.945	0.0842	0.0384	29	-15.89	126.99	2
127	Histidine (%)	2188	0.6035	0.0070	0.6921	0.0530	0.0116	30	-1.67	127.05	0
127	Histidine (%)	0889	0.6075	0.0150	0.6921	0.0530	0.0116	30	-1.60	127.99	0
127	Histidine (%)	0353	0.6200	0.0000	0.6921	0.0530	0.0116	30	-1.36	127.00	0
127	Histidine (%)	0918	0.6284	0.0035	0.6921	0.0530	0.0116	30	-1.20	127.00	0
127	Histidine (%)	0676	0.6350	0.0100	0.6921	0.0530	0.0116	30	-1.08	127.05	0
127	Histidine (%)	0626	0.6435	0.0210	0.6921	0.0530	0.0116	30	-0.92	127.05	0
127	Histidine (%)	0148	0.6550	0.0500	0.6921	0.0530	0.0116	30	-0.70	127.05	0
127	Histidine (%)	0859	0.6575	0.0101	0.6921	0.0530	0.0116	30	-0.65	127.00	0
127	Histidine (%)	0675	0.6650	0.0100	0.6921	0.0530	0.0116	30	-0.51	127.00	0
127	Histidine (%)	0968	0.6800	0.0000	0.6921	0.0530	0.0116	30	-0.23	127.00	0
127	Histidine (%)	0407	0.6815	0.0090	0.6921	0.0530	0.0116	30	-0.20	127.05	0
127	Histidine (%)	0939	0.6850	0.0100	0.6921	0.0530	0.0116	30	-0.13	127.00	0
127	Histidine (%)	0098	0.6860	0.0080	0.6921	0.0530	0.0116	30	-0.11	127.02	0
127	Histidine (%)	0571	0.6875	0.0070	0.6921	0.0530	0.0116	30	-0.09	127.00	0
127	Histidine (%)	2059	0.6875	0.0030	0.6921	0.0530	0.0116	30	-0.09	127.00	0
127	Histidine (%)	0870	0.6918	0.0080	0.6921	0.0530	0.0116	30	0.00	127.00	0
127	Histidine (%)	0910	0.6950	0.0300	0.6921	0.0530	0.0116	30	0.06	127.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
127	Histidine (%)	0872	0.6985	0.0030	0.6921	0.0530	0.0116	30	0.12	127.00	0
127	Histidine (%)	0227	0.7000	0.0200	0.6921	0.0530	0.0116	30	0.15	127.99	0
127	Histidine (%)	0610	0.7050	0.0100	0.6921	0.0530	0.0116	30	0.24	127.05	0
127	Histidine (%)	0644	0.7065	0.0030	0.6921	0.0530	0.0116	30	0.27	127.00	0
127	Histidine (%)	0354	0.7155	0.0010	0.6921	0.0530	0.0116	30	0.44	127.00	0
127	Histidine (%)	0682	0.7260	0.0000	0.6921	0.0530	0.0116	30	0.64	127.00	0
127	Histidine (%)	0878	0.7335	0.0010	0.6921	0.0530	0.0116	30	0.78	127.00	0
127	Histidine (%)	0941	0.7345	0.0430	0.6921	0.0530	0.0116	30	0.80	127.00	0
127	Histidine (%)	0652	0.7400	0.0000	0.6921	0.0530	0.0116	30	0.90	127.00	0
127	Histidine (%)	0868	0.7640	0.0300	0.6921	0.0530	0.0116	30	1.36	127.00	0
127	Histidine (%)	0619	0.7900	0.0220	0.6921	0.0530	0.0116	30	1.85	127.00	0
127	Histidine (%)	2115	0.8340	0.0020	0.6921	0.0530	0.0116	30	2.68	127.99	0
127	Histidine (%)	0226	0.8550	0.0100	0.6921	0.0530	0.0116	30	3.07	127.00	0
128	Isoleucine (%)	0353	0.8350	0.0100	1.236	0.0594	0.0258	29	-6.74	128.00	0
128	Isoleucine (%)	0675	1.045	0.0300	1.236	0.0594	0.0258	29	-3.21	128.00	0
128	Isoleucine (%)	0676	1.110	0.0800	1.236	0.0594	0.0258	29	-2.11	128.05	0
128	Isoleucine (%)	0918	1.112	0.0161	1.236	0.0594	0.0258	29	-2.09	128.00	0
128	Isoleucine (%)	0626	1.170	0.0390	1.236	0.0594	0.0258	29	-1.11	128.05	0
128	Isoleucine (%)	0868	1.195	0.0150	1.236	0.0594	0.0258	29	-0.69	128.00	0
128	Isoleucine (%)	0226	1.200	0.0600	1.236	0.0594	0.0258	29	-0.60	128.00	0
128	Isoleucine (%)	0652	1.200	0.0200	1.236	0.0594	0.0258	29	-0.60	128.00	0
128	Isoleucine (%)	0870	1.221	0.0214	1.236	0.0594	0.0258	29	-0.24	128.00	0
128	Isoleucine (%)	0407	1.222	0.0740	1.236	0.0594	0.0258	29	-0.23	128.05	0
128	Isoleucine (%)	0878	1.224	0.0040	1.236	0.0594	0.0258	29	-0.20	128.00	0
128	Isoleucine (%)	0968	1.225	0.0100	1.236	0.0594	0.0258	29	-0.18	128.00	0
128	Isoleucine (%)	0682	1.227	0.0000	1.236	0.0594	0.0258	29	-0.15	128.00	0
128	Isoleucine (%)	0859	1.234	0.0070	1.236	0.0594	0.0258	29	-0.04	128.00	0
128	Isoleucine (%)	0939	1.235	0.0100	1.236	0.0594	0.0258	29	-0.01	128.00	0
128	Isoleucine (%)	0227	1.240	0.0200	1.236	0.0594	0.0258	29	0.07	128.99	0
128	Isoleucine (%)	0098	1.243	0.0090	1.236	0.0594	0.0258	29	0.12	128.02	0
128	Isoleucine (%)	0872	1.250	0.0390	1.236	0.0594	0.0258	29	0.23	128.00	0
128	Isoleucine (%)	0354	1.254	0.0010	1.236	0.0594	0.0258	29	0.30	128.00	0
128	Isoleucine (%)	0571	1.260	0.0190	1.236	0.0594	0.0258	29	0.40	128.00	0
128	Isoleucine (%)	0941	1.262	0.0720	1.236	0.0594	0.0258	29	0.44	128.00	0
128	Isoleucine (%)	2059	1.270	0.0120	1.236	0.0594	0.0258	29	0.58	128.00	0
128	Isoleucine (%)	0910	1.280	0.0200	1.236	0.0594	0.0258	29	0.75	128.00	0
128	Isoleucine (%)	0610	1.290	0.0200	1.236	0.0594	0.0258	29	0.92	128.05	0
128	Isoleucine (%)	0148	1.300	0.1000	1.236	0.0594	0.0258	29	1.08	128.05	0
128	Isoleucine (%)	2115	1.305	0.0290	1.236	0.0594	0.0258	29	1.16	128.99	0
128	Isoleucine (%)	0644	1.310	0.0010	1.236	0.0594	0.0258	29	1.24	128.00	0
128	Isoleucine (%)	2188	1.310	0.0000	1.236	0.0594	0.0258	29	1.25	128.05	0
128	Isoleucine (%)	0889	1.325	0.0100	1.236	0.0594	0.0258	29	1.50	128.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
128	Isoleucine (%)	0619	1.075	0.1500	1.236	0.0594	0.0258	29	-2.70	128.00	1
129	Leucine (%)	0676	2.585	0.1900	2.899	0.1170	0.0523	30	-2.68	129.05	0
129	Leucine (%)	0353	2.700	0.0400	2.899	0.1170	0.0523	30	-1.70	129.00	0
129	Leucine (%)	0918	2.788	0.0032	2.899	0.1170	0.0523	30	-0.94	129.00	0
129	Leucine (%)	0652	2.805	0.0100	2.899	0.1170	0.0523	30	-0.80	129.00	0
129	Leucine (%)	0859	2.810	0.0170	2.899	0.1170	0.0523	30	-0.76	129.00	0
129	Leucine (%)	0619	2.810	0.0400	2.899	0.1170	0.0523	30	-0.76	129.00	0
129	Leucine (%)	2115	2.812	0.0310	2.899	0.1170	0.0523	30	-0.75	129.99	0
129	Leucine (%)	0148	2.815	0.2100	2.899	0.1170	0.0523	30	-0.72	129.05	0
129	Leucine (%)	0682	2.816	0.0000	2.899	0.1170	0.0523	30	-0.71	129.00	0
129	Leucine (%)	0870	2.825	0.0886	2.899	0.1170	0.0523	30	-0.64	129.00	0
129	Leucine (%)	2188	2.833	0.0050	2.899	0.1170	0.0523	30	-0.57	129.05	0
129	Leucine (%)	2059	2.863	0.0070	2.899	0.1170	0.0523	30	-0.31	129.00	0
129	Leucine (%)	0941	2.866	0.1440	2.899	0.1170	0.0523	30	-0.28	129.00	0
129	Leucine (%)	0098	2.874	0.0230	2.899	0.1170	0.0523	30	-0.22	129.02	0
129	Leucine (%)	0571	2.875	0.0230	2.899	0.1170	0.0523	30	-0.21	129.00	0
129	Leucine (%)	0626	2.879	0.2230	2.899	0.1170	0.0523	30	-0.17	129.05	0
129	Leucine (%)	0878	2.891	0.0150	2.899	0.1170	0.0523	30	-0.07	129.00	0
129	Leucine (%)	0939	2.905	0.0100	2.899	0.1170	0.0523	30	0.05	129.00	0
129	Leucine (%)	0872	2.929	0.0090	2.899	0.1170	0.0523	30	0.25	129.00	0
129	Leucine (%)	0644	2.931	0.0020	2.899	0.1170	0.0523	30	0.27	129.00	0
129	Leucine (%)	0910	2.945	0.0700	2.899	0.1170	0.0523	30	0.39	129.00	0
129	Leucine (%)	0968	2.965	0.0100	2.899	0.1170	0.0523	30	0.57	129.00	0
129	Leucine (%)	0227	2.965	0.0700	2.899	0.1170	0.0523	30	0.57	129.99	0
129	Leucine (%)	0610	2.970	0.0200	2.899	0.1170	0.0523	30	0.61	129.05	0
129	Leucine (%)	0868	3.010	0.0030	2.899	0.1170	0.0523	30	0.95	129.00	0
129	Leucine (%)	0675	3.045	0.0700	2.899	0.1170	0.0523	30	1.25	129.00	0
129	Leucine (%)	0354	3.085	0.0070	2.899	0.1170	0.0523	30	1.59	129.00	0
129	Leucine (%)	0889	3.093	0.0150	2.899	0.1170	0.0523	30	1.65	129.99	0
129	Leucine (%)	0226	3.230	0.1200	2.899	0.1170	0.0523	30	2.83	129.00	0
129	Leucine (%)	0407	3.532	0.0920	2.899	0.1170	0.0523	30	5.41	129.05	0
130	L-Lysine (%)	2053	1.200	0.0000	1.495	0.0832	0.0273	32	-3.55	130.99	0
130	L-Lysine (%)	0676	1.320	0.1000	1.495	0.0832	0.0273	32	-2.11	130.05	0
130	L-Lysine (%)	2188	1.386	0.0010	1.495	0.0832	0.0273	32	-1.32	130.05	0
130	L-Lysine (%)	0619	1.400	0.0600	1.495	0.0832	0.0273	32	-1.14	130.00	0
130	L-Lysine (%)	0918	1.409	0.0105	1.495	0.0832	0.0273	32	-1.03	130.00	0
130	L-Lysine (%)	0860	1.410	0.0000	1.495	0.0832	0.0273	32	-1.02	130.00	0
130	L-Lysine (%)	0682	1.418	0.0000	1.495	0.0832	0.0273	32	-0.93	130.00	0
130	L-Lysine (%)	0652	1.450	0.0200	1.495	0.0832	0.0273	32	-0.54	130.00	0
130	L-Lysine (%)	0353	1.455	0.0500	1.495	0.0832	0.0273	32	-0.48	130.00	0
130	L-Lysine (%)	0675	1.460	0.0400	1.495	0.0832	0.0273	32	-0.42	130.00	0
130	L-Lysine (%)	0939	1.470	0.0200	1.495	0.0832	0.0273	32	-0.30	130.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
130	L-Lysine (%)	0968	1.470	0.0400	1.495	0.0832	0.0273	32	-0.30	130.00	0
130	L-Lysine (%)	0859	1.478	0.0097	1.495	0.0832	0.0273	32	-0.21	130.00	0
130	L-Lysine (%)	0848	1.485	0.0100	1.495	0.0832	0.0273	32	-0.12	130.00	0
130	L-Lysine (%)	0626	1.486	0.1020	1.495	0.0832	0.0273	32	-0.11	130.05	0
130	L-Lysine (%)	0872	1.487	0.0070	1.495	0.0832	0.0273	32	-0.11	130.00	0
130	L-Lysine (%)	0941	1.499	0.0720	1.495	0.0832	0.0273	32	0.05	130.00	0
130	L-Lysine (%)	0878	1.502	0.0050	1.495	0.0832	0.0273	32	0.08	130.00	0
130	L-Lysine (%)	0354	1.502	0.0020	1.495	0.0832	0.0273	32	0.08	130.00	0
130	L-Lysine (%)	0910	1.510	0.0600	1.495	0.0832	0.0273	32	0.18	130.00	0
130	L-Lysine (%)	0610	1.515	0.0100	1.495	0.0832	0.0273	32	0.24	130.05	0
130	L-Lysine (%)	0644	1.517	0.0010	1.495	0.0832	0.0273	32	0.26	130.00	0
130	L-Lysine (%)	0571	1.519	0.0230	1.495	0.0832	0.0273	32	0.28	130.00	0
130	L-Lysine (%)	2059	1.529	0.0040	1.495	0.0832	0.0273	32	0.41	130.00	0
130	L-Lysine (%)	0868	1.539	0.0410	1.495	0.0832	0.0273	32	0.52	130.00	0
130	L-Lysine (%)	0098	1.575	0.0320	1.495	0.0832	0.0273	32	0.96	130.02	0
130	L-Lysine (%)	0870	1.578	0.0234	1.495	0.0832	0.0273	32	0.99	130.00	0
130	L-Lysine (%)	0148	1.585	0.0100	1.495	0.0832	0.0273	32	1.08	130.05	0
130	L-Lysine (%)	2115	1.616	0.0300	1.495	0.0832	0.0273	32	1.45	130.99	0
130	L-Lysine (%)	0227	1.685	0.0300	1.495	0.0832	0.0273	32	2.28	130.99	0
130	L-Lysine (%)	0226	1.770	0.0400	1.495	0.0832	0.0273	32	3.30	130.00	0
130	L-Lysine (%)	0889	2.105	0.0200	1.495	0.0832	0.0273	32	7.33	130.99	0
130	L-Lysine (%)	0407	1.470	0.1950	1.495	0.0832	0.0273	32	-0.31	130.05	1
131	Methionine (%)	0675	0.3050	0.0300	0.5753	0.0368	0.0109	31	-7.34	131.00	0
131	Methionine (%)	0354	0.5275	0.0030	0.5753	0.0368	0.0109	31	-1.30	131.00	0
131	Methionine (%)	0941	0.5340	0.0080	0.5753	0.0368	0.0109	31	-1.12	131.00	0
131	Methionine (%)	0682	0.5370	0.0000	0.5753	0.0368	0.0109	31	-1.04	131.00	0
131	Methionine (%)	0859	0.5396	0.0055	0.5753	0.0368	0.0109	31	-0.97	131.00	0
131	Methionine (%)	0918	0.5400	0.0000	0.5753	0.0368	0.0109	31	-0.96	131.00	0
131	Methionine (%)	0939	0.5400	0.0200	0.5753	0.0368	0.0109	31	-0.96	131.00	0
131	Methionine (%)	2115	0.5435	0.0230	0.5753	0.0368	0.0109	31	-0.86	131.99	0
131	Methionine (%)	0619	0.5545	0.0070	0.5753	0.0368	0.0109	31	-0.56	131.00	0
131	Methionine (%)	0652	0.5550	0.0100	0.5753	0.0368	0.0109	31	-0.55	131.00	0
131	Methionine (%)	0968	0.5600	0.0040	0.5753	0.0368	0.0109	31	-0.42	131.00	0
131	Methionine (%)	2059	0.5620	0.0060	0.5753	0.0368	0.0109	31	-0.36	131.00	0
131	Methionine (%)	0878	0.5635	0.0010	0.5753	0.0368	0.0109	31	-0.32	131.00	0
131	Methionine (%)	0626	0.5710	0.0500	0.5753	0.0368	0.0109	31	-0.12	131.05	0
131	Methionine (%)	2188	0.5725	0.0050	0.5753	0.0368	0.0109	31	-0.08	131.05	0
131	Methionine (%)	0098	0.5740	0.0040	0.5753	0.0368	0.0109	31	-0.03	131.02	0
131	Methionine (%)	0872	0.5750	0.0000	0.5753	0.0368	0.0109	31	-0.01	131.00	0
131	Methionine (%)	0610	0.5800	0.0000	0.5753	0.0368	0.0109	31	0.13	131.05	0
131	Methionine (%)	0571	0.5805	0.0110	0.5753	0.0368	0.0109	31	0.14	131.00	0
131	Methionine (%)	0868	0.5850	0.0160	0.5753	0.0368	0.0109	31	0.26	131.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
131	Methionine (%)	0870	0.5933	0.0011	0.5753	0.0368	0.0109	31	0.49	131.00	0
131	Methionine (%)	0644	0.5950	0.0040	0.5753	0.0368	0.0109	31	0.54	131.00	0
131	Methionine (%)	0848	0.5950	0.0100	0.5753	0.0368	0.0109	31	0.54	131.00	0
131	Methionine (%)	0910	0.5950	0.0300	0.5753	0.0368	0.0109	31	0.54	131.00	0
131	Methionine (%)	0407	0.5950	0.0440	0.5753	0.0368	0.0109	31	0.54	131.05	0
131	Methionine (%)	0227	0.6100	0.0000	0.5753	0.0368	0.0109	31	0.94	131.99	0
131	Methionine (%)	0676	0.6140	0.0060	0.5753	0.0368	0.0109	31	1.05	131.05	0
131	Methionine (%)	0226	0.6500	0.0200	0.5753	0.0368	0.0109	31	2.03	131.00	0
131	Methionine (%)	0353	0.7500	0.0000	0.5753	0.0368	0.0109	31	4.74	131.00	0
131	Methionine (%)	0889	0.7650	0.0100	0.5753	0.0368	0.0109	31	5.15	131.99	0
131	Methionine (%)	2053	0.7950	0.0100	0.5753	0.0368	0.0109	31	5.97	131.99	0
131	Methionine (%)	0148	0.4700	0.0600	0.5753	0.0368	0.0109	31	-2.86	131.05	1
132	Phenylalanine (%)	0675	1.335	0.0100	1.468	0.0796	0.0279	29	-1.67	132.00	0
132	Phenylalanine (%)	0682	1.354	0.0000	1.468	0.0796	0.0279	29	-1.44	132.00	0
132	Phenylalanine (%)	0676	1.355	0.0700	1.468	0.0796	0.0279	29	-1.42	132.05	0
132	Phenylalanine (%)	0619	1.365	0.0700	1.468	0.0796	0.0279	29	-1.30	132.00	0
132	Phenylalanine (%)	0889	1.373	0.0050	1.468	0.0796	0.0279	29	-1.20	132.99	0
132	Phenylalanine (%)	0652	1.395	0.0500	1.468	0.0796	0.0279	29	-0.92	132.00	0
132	Phenylalanine (%)	0626	1.421	0.1110	1.468	0.0796	0.0279	29	-0.60	132.05	0
132	Phenylalanine (%)	0939	1.430	0.0200	1.468	0.0796	0.0279	29	-0.48	132.00	0
132	Phenylalanine (%)	2188	1.433	0.0050	1.468	0.0796	0.0279	29	-0.45	132.05	0
132	Phenylalanine (%)	2059	1.440	0.0110	1.468	0.0796	0.0279	29	-0.36	132.00	0
132	Phenylalanine (%)	0878	1.460	0.0010	1.468	0.0796	0.0279	29	-0.11	132.00	0
132	Phenylalanine (%)	0353	1.460	0.0200	1.468	0.0796	0.0279	29	-0.10	132.00	0
132	Phenylalanine (%)	0098	1.469	0.0150	1.468	0.0796	0.0279	29	0.00	132.02	0
132	Phenylalanine (%)	0571	1.470	0.0270	1.468	0.0796	0.0279	29	0.02	132.00	0
132	Phenylalanine (%)	0968	1.470	0.0200	1.468	0.0796	0.0279	29	0.02	132.00	0
132	Phenylalanine (%)	2115	1.473	0.0240	1.468	0.0796	0.0279	29	0.06	132.99	0
132	Phenylalanine (%)	0870	1.482	0.0175	1.468	0.0796	0.0279	29	0.18	132.00	0
132	Phenylalanine (%)	0872	1.487	0.0050	1.468	0.0796	0.0279	29	0.23	132.00	0
132	Phenylalanine (%)	0227	1.490	0.0000	1.468	0.0796	0.0279	29	0.27	132.99	0
132	Phenylalanine (%)	0859	1.493	0.0095	1.468	0.0796	0.0279	29	0.31	132.00	0
132	Phenylalanine (%)	0941	1.500	0.0510	1.468	0.0796	0.0279	29	0.39	132.00	0
132	Phenylalanine (%)	0910	1.505	0.0500	1.468	0.0796	0.0279	29	0.46	132.00	0
132	Phenylalanine (%)	0644	1.519	0.0120	1.468	0.0796	0.0279	29	0.64	132.00	0
132	Phenylalanine (%)	0354	1.520	0.0010	1.468	0.0796	0.0279	29	0.64	132.00	0
132	Phenylalanine (%)	0868	1.548	0.0490	1.468	0.0796	0.0279	29	1.00	132.00	0
132	Phenylalanine (%)	0610	1.560	0.0200	1.468	0.0796	0.0279	29	1.15	132.05	0
132	Phenylalanine (%)	0226	1.645	0.0700	1.468	0.0796	0.0279	29	2.22	132.00	0
132	Phenylalanine (%)	0918	1.729	0.0107	1.468	0.0796	0.0279	29	3.27	132.00	0
132	Phenylalanine (%)	0407	1.889	0.0550	1.468	0.0796	0.0279	29	5.28	132.05	0
132	Phenylalanine (%)	0148	1.350	0.1600	1.468	0.0796	0.0279	29	-1.49	132.05	1

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
133	Proline (%)	0859	1.194	0.0141	2.153	0.0810	0.0624	29	-11.84	133.00	0
133	Proline (%)	0918	1.797	0.0309	2.153	0.0810	0.0624	29	-4.39	133.00	0
133	Proline (%)	0682	1.930	0.0000	2.153	0.0810	0.0624	29	-2.75	133.00	0
133	Proline (%)	0676	1.945	0.0500	2.153	0.0810	0.0624	29	-2.56	133.05	0
133	Proline (%)	2115	2.082	0.0040	2.153	0.0810	0.0624	29	-0.87	133.99	0
133	Proline (%)	0889	2.098	0.0150	2.153	0.0810	0.0624	29	-0.68	133.99	0
133	Proline (%)	0968	2.100	0.0200	2.153	0.0810	0.0624	29	-0.65	133.00	0
133	Proline (%)	0870	2.107	0.1162	2.153	0.0810	0.0624	29	-0.57	133.00	0
133	Proline (%)	0626	2.109	0.0080	2.153	0.0810	0.0624	29	-0.54	133.05	0
133	Proline (%)	0148	2.110	0.1600	2.153	0.0810	0.0624	29	-0.53	133.05	0
133	Proline (%)	0353	2.125	0.0500	2.153	0.0810	0.0624	29	-0.34	133.00	0
133	Proline (%)	0910	2.130	0.1000	2.153	0.0810	0.0624	29	-0.28	133.00	0
133	Proline (%)	0227	2.160	0.0400	2.153	0.0810	0.0624	29	0.09	133.99	0
133	Proline (%)	0571	2.161	0.0240	2.153	0.0810	0.0624	29	0.10	133.00	0
133	Proline (%)	0878	2.166	0.0010	2.153	0.0810	0.0624	29	0.16	133.00	0
133	Proline (%)	0644	2.168	0.0330	2.153	0.0810	0.0624	29	0.18	133.00	0
133	Proline (%)	0652	2.175	0.0700	2.153	0.0810	0.0624	29	0.28	133.00	0
133	Proline (%)	0941	2.180	0.1670	2.153	0.0810	0.0624	29	0.33	133.00	0
133	Proline (%)	0226	2.180	0.2600	2.153	0.0810	0.0624	29	0.34	133.00	0
133	Proline (%)	0872	2.184	0.0180	2.153	0.0810	0.0624	29	0.39	133.00	0
133	Proline (%)	0868	2.187	0.0990	2.153	0.0810	0.0624	29	0.42	133.00	0
133	Proline (%)	2059	2.190	0.0060	2.153	0.0810	0.0624	29	0.46	133.00	0
133	Proline (%)	0610	2.210	0.0200	2.153	0.0810	0.0624	29	0.71	133.05	0
133	Proline (%)	0939	2.215	0.0300	2.153	0.0810	0.0624	29	0.77	133.00	0
133	Proline (%)	0407	2.234	0.2560	2.153	0.0810	0.0624	29	1.00	133.05	0
133	Proline (%)	0619	2.235	0.1300	2.153	0.0810	0.0624	29	1.02	133.00	0
133	Proline (%)	2188	2.252	0.0030	2.153	0.0810	0.0624	29	1.22	133.05	0
133	Proline (%)	0354	2.331	0.0030	2.153	0.0810	0.0624	29	2.20	133.00	0
133	Proline (%)	0675	2.690	0.0800	2.153	0.0810	0.0624	29	6.64	133.00	0
134	Serine (%)	0407	1.051	0.1120	1.442	0.0873	0.0378	30	-4.48	134.05	0
134	Serine (%)	0098	1.320	0.0800	1.442	0.0873	0.0378	30	-1.40	134.02	0
134	Serine (%)	0676	1.320	0.1000	1.442	0.0873	0.0378	30	-1.40	134.05	0
134	Serine (%)	0675	1.355	0.0100	1.442	0.0873	0.0378	30	-1.00	134.00	0
134	Serine (%)	0148	1.360	0.1000	1.442	0.0873	0.0378	30	-0.94	134.05	0
134	Serine (%)	2059	1.362	0.0160	1.442	0.0873	0.0378	30	-0.92	134.00	0
134	Serine (%)	0644	1.376	0.0110	1.442	0.0873	0.0378	30	-0.76	134.00	0
134	Serine (%)	0610	1.380	0.0000	1.442	0.0873	0.0378	30	-0.71	134.05	0
134	Serine (%)	0626	1.384	0.0780	1.442	0.0873	0.0378	30	-0.66	134.05	0
134	Serine (%)	0918	1.389	0.0085	1.442	0.0873	0.0378	30	-0.61	134.00	0
134	Serine (%)	0941	1.399	0.0680	1.442	0.0873	0.0378	30	-0.49	134.00	0
134	Serine (%)	2188	1.420	0.0000	1.442	0.0873	0.0378	30	-0.25	134.05	0
134	Serine (%)	0571	1.430	0.0280	1.442	0.0873	0.0378	30	-0.14	134.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
134	Serine (%)	0968	1.430	0.0200	1.442	0.0873	0.0378	30	-0.14	134.00	0
134	Serine (%)	0878	1.442	0.0020	1.442	0.0873	0.0378	30	0.00	134.00	0
134	Serine (%)	0652	1.445	0.0300	1.442	0.0873	0.0378	30	0.03	134.00	0
134	Serine (%)	0859	1.447	0.0323	1.442	0.0873	0.0378	30	0.06	134.00	0
134	Serine (%)	0872	1.461	0.0620	1.442	0.0873	0.0378	30	0.22	134.00	0
134	Serine (%)	2115	1.465	0.0440	1.442	0.0873	0.0378	30	0.26	134.99	0
134	Serine (%)	0910	1.470	0.0800	1.442	0.0873	0.0378	30	0.32	134.00	0
134	Serine (%)	0353	1.475	0.0100	1.442	0.0873	0.0378	30	0.38	134.00	0
134	Serine (%)	0227	1.485	0.0100	1.442	0.0873	0.0378	30	0.49	134.99	0
134	Serine (%)	0868	1.523	0.0340	1.442	0.0873	0.0378	30	0.93	134.00	0
134	Serine (%)	0682	1.526	0.0000	1.442	0.0873	0.0378	30	0.96	134.00	0
134	Serine (%)	0354	1.530	0.0100	1.442	0.0873	0.0378	30	1.01	134.00	0
134	Serine (%)	0619	1.530	0.0400	1.442	0.0873	0.0378	30	1.01	134.00	0
134	Serine (%)	0939	1.535	0.0500	1.442	0.0873	0.0378	30	1.06	134.00	0
134	Serine (%)	0870	1.545	0.0841	1.442	0.0873	0.0378	30	1.18	134.00	0
134	Serine (%)	0889	1.578	0.0050	1.442	0.0873	0.0378	30	1.55	134.99	0
134	Serine (%)	0226	1.695	0.0100	1.442	0.0873	0.0378	30	2.90	134.00	0
135	Threonine (%)	0676	1.040	0.0800	1.155	0.0424	0.0215	30	-2.72	135.05	0
135	Threonine (%)	0407	1.046	0.0220	1.155	0.0424	0.0215	30	-2.58	135.05	0
135	Threonine (%)	0353	1.095	0.0300	1.155	0.0424	0.0215	30	-1.42	135.00	0
135	Threonine (%)	0148	1.105	0.0700	1.155	0.0424	0.0215	30	-1.18	135.05	0
135	Threonine (%)	0626	1.108	0.0410	1.155	0.0424	0.0215	30	-1.13	135.05	0
135	Threonine (%)	2059	1.113	0.0010	1.155	0.0424	0.0215	30	-1.01	135.00	0
135	Threonine (%)	0859	1.119	0.0147	1.155	0.0424	0.0215	30	-0.85	135.00	0
135	Threonine (%)	0652	1.120	0.0000	1.155	0.0424	0.0215	30	-0.83	135.00	0
135	Threonine (%)	0675	1.135	0.0100	1.155	0.0424	0.0215	30	-0.48	135.00	0
135	Threonine (%)	0941	1.140	0.0550	1.155	0.0424	0.0215	30	-0.37	135.00	0
135	Threonine (%)	0682	1.152	0.0000	1.155	0.0424	0.0215	30	-0.07	135.00	0
135	Threonine (%)	2188	1.152	0.0040	1.155	0.0424	0.0215	30	-0.07	135.05	0
135	Threonine (%)	0571	1.153	0.0260	1.155	0.0424	0.0215	30	-0.05	135.00	0
135	Threonine (%)	0644	1.156	0.0130	1.155	0.0424	0.0215	30	0.01	135.00	0
135	Threonine (%)	0878	1.160	0.0050	1.155	0.0424	0.0215	30	0.10	135.00	0
135	Threonine (%)	0098	1.160	0.0080	1.155	0.0424	0.0215	30	0.11	135.02	0
135	Threonine (%)	0918	1.163	0.0216	1.155	0.0424	0.0215	30	0.19	135.00	0
135	Threonine (%)	0939	1.165	0.0100	1.155	0.0424	0.0215	30	0.23	135.00	0
135	Threonine (%)	0968	1.165	0.0100	1.155	0.0424	0.0215	30	0.23	135.00	0
135	Threonine (%)	0354	1.167	0.0020	1.155	0.0424	0.0215	30	0.28	135.00	0
135	Threonine (%)	0872	1.170	0.0400	1.155	0.0424	0.0215	30	0.35	135.00	0
135	Threonine (%)	0619	1.175	0.0100	1.155	0.0424	0.0215	30	0.47	135.00	0
135	Threonine (%)	0610	1.180	0.0000	1.155	0.0424	0.0215	30	0.59	135.05	0
135	Threonine (%)	2115	1.185	0.0370	1.155	0.0424	0.0215	30	0.69	135.99	0
135	Threonine (%)	0868	1.187	0.0130	1.155	0.0424	0.0215	30	0.74	135.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
135	Threonine (%)	0227	1.195	0.0300	1.155	0.0424	0.0215	30	0.94	135.99	0
135	Threonine (%)	0910	1.205	0.0500	1.155	0.0424	0.0215	30	1.18	135.00	0
135	Threonine (%)	0889	1.213	0.0050	1.155	0.0424	0.0215	30	1.35	135.99	0
135	Threonine (%)	0870	1.220	0.0164	1.155	0.0424	0.0215	30	1.53	135.00	0
135	Threonine (%)	0226	1.330	0.0200	1.155	0.0424	0.0215	30	4.13	135.00	0
136	Tryptophan (%)	0682	0.1980	0.0000	0.3175	0.0388	0.0073	21	-3.08	136.00	0
136	Tryptophan (%)	0407	0.2380	0.0060	0.3175	0.0388	0.0073	21	-2.05	136.05	0
136	Tryptophan (%)	0910	0.2750	0.0100	0.3175	0.0388	0.0073	21	-1.10	136.00	0
136	Tryptophan (%)	0918	0.2800	0.0000	0.3175	0.0388	0.0073	21	-0.97	136.00	0
136	Tryptophan (%)	0941	0.2800	0.0100	0.3175	0.0388	0.0073	21	-0.97	136.00	0
136	Tryptophan (%)	0098	0.3020	0.0140	0.3175	0.0388	0.0073	21	-0.40	136.02	0
136	Tryptophan (%)	0872	0.3090	0.0000	0.3175	0.0388	0.0073	21	-0.22	136.03	0
136	Tryptophan (%)	0619	0.3125	0.0150	0.3175	0.0388	0.0073	21	-0.13	136.00	0
136	Tryptophan (%)	2059	0.3125	0.0010	0.3175	0.0388	0.0073	21	-0.13	136.03	0
136	Tryptophan (%)	0148	0.3140	0.0040	0.3175	0.0388	0.0073	21	-0.09	136.03	0
136	Tryptophan (%)	0939	0.3150	0.0100	0.3175	0.0388	0.0073	21	-0.06	136.03	0
136	Tryptophan (%)	0878	0.3155	0.0010	0.3175	0.0388	0.0073	21	-0.05	136.01	0
136	Tryptophan (%)	0644	0.3220	0.0040	0.3175	0.0388	0.0073	21	0.12	136.01	0
136	Tryptophan (%)	0571	0.3250	0.0020	0.3175	0.0388	0.0073	21	0.19	136.03	0
136	Tryptophan (%)	0868	0.3350	0.0060	0.3175	0.0388	0.0073	21	0.45	136.01	0
136	Tryptophan (%)	0227	0.3400	0.0200	0.3175	0.0388	0.0073	21	0.58	136.00	0
136	Tryptophan (%)	0859	0.3439	0.0105	0.3175	0.0388	0.0073	21	0.68	136.03	0
136	Tryptophan (%)	2114	0.3575	0.0060	0.3175	0.0388	0.0073	21	1.03	136.01	0
136	Tryptophan (%)	0870	0.3588	0.0233	0.3175	0.0388	0.0073	21	1.06	136.00	0
136	Tryptophan (%)	0353	0.4050	0.0100	0.3175	0.0388	0.0073	21	2.25	136.03	0
136	Tryptophan (%)	0889	0.5800	0.0000	0.3175	0.0388	0.0073	21	6.76	136.99	0
137	Tyrosine (%)	0889	0.5800	0.0000	1.100	0.0916	0.0349	23	-5.68	137.99	0
137	Tyrosine (%)	0910	0.8700	0.0400	1.100	0.0916	0.0349	23	-2.51	137.00	0
137	Tyrosine (%)	0870	0.9938	0.0272	1.100	0.0916	0.0349	23	-1.16	137.00	0
137	Tyrosine (%)	0227	0.9950	0.0500	1.100	0.0916	0.0349	23	-1.15	137.99	0
137	Tyrosine (%)	0626	1.011	0.1350	1.100	0.0916	0.0349	23	-0.98	137.05	0
137	Tyrosine (%)	0872	1.043	0.0280	1.100	0.0916	0.0349	23	-0.62	137.00	0
137	Tyrosine (%)	0148	1.075	0.0700	1.100	0.0916	0.0349	23	-0.27	137.05	0
137	Tyrosine (%)	0676	1.075	0.0100	1.100	0.0916	0.0349	23	-0.27	137.05	0
137	Tyrosine (%)	2115	1.084	0.0140	1.100	0.0916	0.0349	23	-0.18	137.99	0
137	Tyrosine (%)	0644	1.091	0.0050	1.100	0.0916	0.0349	23	-0.11	137.00	0
137	Tyrosine (%)	0968	1.095	0.0100	1.100	0.0916	0.0349	23	-0.06	137.00	0
137	Tyrosine (%)	0354	1.106	0.0110	1.100	0.0916	0.0349	23	0.06	137.00	0
137	Tyrosine (%)	0918	1.111	0.0050	1.100	0.0916	0.0349	23	0.12	137.00	0
137	Tyrosine (%)	2188	1.112	0.0040	1.100	0.0916	0.0349	23	0.13	137.05	0
137	Tyrosine (%)	0098	1.116	0.0070	1.100	0.0916	0.0349	23	0.17	137.02	0
137	Tyrosine (%)	0682	1.127	0.0000	1.100	0.0916	0.0349	23	0.29	137.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
137	Tyrosine (%)	0675	1.130	0.0200	1.100	0.0916	0.0349	23	0.33	137.00	0
137	Tyrosine (%)	0941	1.158	0.1830	1.100	0.0916	0.0349	23	0.63	137.00	0
137	Tyrosine (%)	0226	1.170	0.0200	1.100	0.0916	0.0349	23	0.76	137.00	0
137	Tyrosine (%)	0939	1.205	0.1300	1.100	0.0916	0.0349	23	1.15	137.00	0
137	Tyrosine (%)	2059	1.210	0.0040	1.100	0.0916	0.0349	23	1.20	137.00	0
137	Tyrosine (%)	0610	1.235	0.0100	1.100	0.0916	0.0349	23	1.47	137.05	0
137	Tyrosine (%)	0353	1.260	0.0200	1.100	0.0916	0.0349	23	1.75	137.00	0
137	Tyrosine (%)	0407	0.7600	1.176	1.100	0.0916	0.0349	23	-3.71	137.05	1
138	Valine (%)	0889	1.043	0.0150	1.397	0.1125	0.0316	29	-3.15	138.99	0
138	Valine (%)	2188	1.061	0.0020	1.397	0.1125	0.0316	29	-2.99	138.05	0
138	Valine (%)	0353	1.105	0.0300	1.397	0.1125	0.0316	29	-2.60	138.00	0
138	Valine (%)	0675	1.145	0.0100	1.397	0.1125	0.0316	29	-2.24	138.00	0
138	Valine (%)	0226	1.200	0.0600	1.397	0.1125	0.0316	29	-1.75	138.00	0
138	Valine (%)	0619	1.245	0.1500	1.397	0.1125	0.0316	29	-1.35	138.00	0
138	Valine (%)	0941	1.339	0.0900	1.397	0.1125	0.0316	29	-0.52	138.00	0
138	Valine (%)	0676	1.355	0.0500	1.397	0.1125	0.0316	29	-0.37	138.05	0
138	Valine (%)	0870	1.378	0.0554	1.397	0.1125	0.0316	29	-0.17	138.00	0
138	Valine (%)	0918	1.380	0.0130	1.397	0.1125	0.0316	29	-0.15	138.00	0
138	Valine (%)	0652	1.390	0.0600	1.397	0.1125	0.0316	29	-0.06	138.00	0
138	Valine (%)	0859	1.409	0.0111	1.397	0.1125	0.0316	29	0.11	138.00	0
138	Valine (%)	0868	1.409	0.0160	1.397	0.1125	0.0316	29	0.11	138.00	0
138	Valine (%)	0626	1.419	0.0630	1.397	0.1125	0.0316	29	0.19	138.05	0
138	Valine (%)	0682	1.421	0.0000	1.397	0.1125	0.0316	29	0.21	138.00	0
138	Valine (%)	0968	1.425	0.0100	1.397	0.1125	0.0316	29	0.25	138.00	0
138	Valine (%)	0910	1.435	0.0100	1.397	0.1125	0.0316	29	0.34	138.00	0
138	Valine (%)	0354	1.441	0.0050	1.397	0.1125	0.0316	29	0.39	138.00	0
138	Valine (%)	0878	1.445	0.0060	1.397	0.1125	0.0316	29	0.43	138.00	0
138	Valine (%)	0939	1.445	0.0100	1.397	0.1125	0.0316	29	0.43	138.00	0
138	Valine (%)	2115	1.454	0.0210	1.397	0.1125	0.0316	29	0.50	138.99	0
138	Valine (%)	0571	1.458	0.0240	1.397	0.1125	0.0316	29	0.54	138.00	0
138	Valine (%)	0872	1.465	0.0460	1.397	0.1125	0.0316	29	0.60	138.00	0
138	Valine (%)	0148	1.475	0.0700	1.397	0.1125	0.0316	29	0.69	138.05	0
138	Valine (%)	2059	1.484	0.0080	1.397	0.1125	0.0316	29	0.77	138.00	0
138	Valine (%)	0098	1.489	0.0220	1.397	0.1125	0.0316	29	0.82	138.02	0
138	Valine (%)	0227	1.510	0.0400	1.397	0.1125	0.0316	29	1.00	138.99	0
138	Valine (%)	0644	1.536	0.0180	1.397	0.1125	0.0316	29	1.24	138.00	0
138	Valine (%)	0610	1.610	0.0000	1.397	0.1125	0.0316	29	1.89	138.05	0
138	Valine (%)	0407	1.446	0.3850	1.397	0.1125	0.0316	29	0.43	138.05	1
139	Taurine (%)	0941	0.1220	0.0000	0.1644	0.0421	0.0046	13	-1.01	139.00	0
139	Taurine (%)	0644	0.1295	0.0010	0.1644	0.0421	0.0046	13	-0.83	139.00	0
139	Taurine (%)	0939	0.1300	0.0000	0.1644	0.0421	0.0046	13	-0.82	139.99	0
139	Taurine (%)	0148	0.1400	0.0200	0.1644	0.0421	0.0046	13	-0.58	139.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
139	Taurine (%)	0014	0.1420	0.0140	0.1644	0.0421	0.0046	13	-0.53	139.05	0
139	Taurine (%)	0160	0.1450	0.0100	0.1644	0.0421	0.0046	13	-0.46	139.99	0
139	Taurine (%)	0098	0.1520	0.0040	0.1644	0.0421	0.0046	13	-0.30	139.02	0
139	Taurine (%)	0676	0.1520	0.0020	0.1644	0.0421	0.0046	13	-0.30	139.05	0
139	Taurine (%)	0227	0.1700	0.0000	0.1644	0.0421	0.0046	13	0.13	139.99	0
139	Taurine (%)	0889	0.1925	0.0050	0.1644	0.0421	0.0046	13	0.67	139.99	0
139	Taurine (%)	0682	0.2100	0.0000	0.1644	0.0421	0.0046	13	1.08	139.00	0
139	Taurine (%)	0407	0.2250	0.0000	0.1644	0.0421	0.0046	13	1.44	139.05	0
139	Taurine (%)	2115	0.3310	0.0040	0.1644	0.0421	0.0046	13	3.96	139.99	0
160	Fructose (%)	0297	0.1225	0.0070				3		160.10	0
160	Fructose (%)	0096	0.1351	0.0163				3		160.99	0
160	Fructose (%)	0619	0.1480	0.0020				3		160.99	0
160	Fructose (%)	0227	0.3250	0.0100				3		160.99	2
161	Galactose (%)	0297	0.0000	0.0000				0		161.10	4
162	Glucose (%)	0297	0.1430	0.0000				3		162.10	0
162	Glucose (%)	0619	0.1660	0.0000				3		162.99	0
162	Glucose (%)	0227	0.2350	0.0100				3		162.99	0
162	Glucose (%)	0096	< 0.2					3		162.99	5
163	Lactose (%)	0297	0.0000	0.0000				0		163.10	4
163	Lactose (%)	0619	0.0000	0.0000				0		163.99	4
163	Lactose (%)	0227	< 0.15					0		163.99	5
163	Lactose (%)	0096	< 0.4					0		163.99	5
164	Maltose (%)	0297	0.2330	0.0040				2		164.10	0
164	Maltose (%)	0227	0.3000	0.0200				2		164.99	0
164	Maltose (%)	0096	< 0.4					2		164.99	5
165	Sucrose (%)	0297	1.219	0.0280	1.227	0.0072	0.0325	4		165.10	0
165	Sucrose (%)	0096	1.223	0.0718	1.227	0.0072	0.0325	4		165.99	0
165	Sucrose (%)	0619	1.230	0.0200	1.227	0.0072	0.0325	4		165.99	0
165	Sucrose (%)	0227	1.235	0.0100	1.227	0.0072	0.0325	4		165.99	0
166	Raffinose (%)	0619	0.2065	0.0030				3		166.99	0
166	Raffinose (%)	0227	0.2150	0.0100				3		166.99	0
166	Raffinose (%)	0297	0.2210	0.0260				3		166.10	0
167	Stachyose (%)	0297	0.7325	0.0450				3		167.10	0
167	Stachyose (%)	0619	0.7975	0.0110				3		167.99	0
167	Stachyose (%)	0227	0.8150	0.0300				3		167.99	0
400	Water Activity (Units)	0942	0.4490	0.0000	0.4711	0.0168	0.0051	12	-1.31	400.01	0
400	Water Activity (Units)	0843	0.4532	0.0002	0.4711	0.0168	0.0051	12	-1.06	400.01	0
400	Water Activity (Units)	0407	0.4558	0.0001	0.4711	0.0168	0.0051	12	-0.91	400.01	0
400	Water Activity (Units)	0843	0.4565	0.0030	0.4711	0.0168	0.0051	12	-0.87	400.99	0
400	Water Activity (Units)	0589	0.4695	0.0170	0.4711	0.0168	0.0051	12	-0.09	400.01	0
400	Water Activity (Units)	2174	0.4710	0.0060	0.4711	0.0168	0.0051	12	0.00	400.01	0
400	Water Activity (Units)	2268	0.4730	0.0060	0.4711	0.0168	0.0051	12	0.12	400.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
400	Water Activity (Units)	2303	0.4786	0.0090	0.4711	0.0168	0.0051	12	0.45	400.01	0
400	Water Activity (Units)	2073	0.4795	0.0070	0.4711	0.0168	0.0051	12	0.50	400.99	0
400	Water Activity (Units)	2083	0.4850	0.0100	0.4711	0.0168	0.0051	12	0.83	400.99	0
400	Water Activity (Units)	2181	0.4854	0.0008	0.4711	0.0168	0.0051	12	0.85	400.01	0
400	Water Activity (Units)	2109	0.6083	0.0017	0.4711	0.0168	0.0051	12	8.15	400.01	0
400	Water Activity (Units)	2168	0.4550	0.0300	0.4711	0.0168	0.0051	12	-0.95	400.01	1
516	Arsenic, Total (ppm)	0553	0.0472	0.0027	0.0736	0.0383	0.0071	6	-0.69	516.53	0
516	Arsenic, Total (ppm)	0910	0.0500	0.0000	0.0736	0.0383	0.0071	6	-0.62	516.52	0
516	Arsenic, Total (ppm)	0227	0.0555	0.0070	0.0736	0.0383	0.0071	6	-0.47	516.53	0
516	Arsenic, Total (ppm)	2012	0.0595	0.0030	0.0736	0.0383	0.0071	6	-0.37	516.53	0
516	Arsenic, Total (ppm)	0918	0.0985	0.0116	0.0736	0.0383	0.0071	6	0.65	516.53	0
516	Arsenic, Total (ppm)	0407	1.234	0.0180	0.0736	0.0383	0.0071	6	30.28	516.43	0
516	Arsenic, Total (ppm)	0186	< 0.05		0.0736	0.0383	0.0071	6		516.52	5
516	Arsenic, Total (ppm)	2124	< 0.89		0.0736	0.0383	0.0071	6		516.43	5
516	Arsenic, Total (ppm)	0682	< 10		0.0736	0.0383	0.0071	6		516.43	5
518	Cadmium (ppm)	0407	0.0245	0.0010	0.0342	0.0036	0.0028	8	-2.68	518.43	0
518	Cadmium (ppm)	2055	0.0310	0.0020	0.0342	0.0036	0.0028	8	-0.89	518.53	0
518	Cadmium (ppm)	2012	0.0340	0.0020	0.0342	0.0036	0.0028	8	-0.06	518.53	0
518	Cadmium (ppm)	0918	0.0342	0.0022	0.0342	0.0036	0.0028	8	0.00	518.53	0
518	Cadmium (ppm)	0910	0.0350	0.0100	0.0342	0.0036	0.0028	8	0.22	518.52	0
518	Cadmium (ppm)	0227	0.0350	0.0040	0.0342	0.0036	0.0028	8	0.22	518.53	0
518	Cadmium (ppm)	0553	0.0361	0.0014	0.0342	0.0036	0.0028	8	0.52	518.53	0
518	Cadmium (ppm)	0968	0.0460	0.0000	0.0342	0.0036	0.0028	8	3.25	518.43	0
518	Cadmium (ppm)	0407	0.2104	0.3420	0.0342	0.0036	0.0028	8	48.56	518.41	1
518	Cadmium (ppm)	0186	< 0.05		0.0342	0.0036	0.0028	8		518.52	5
518	Cadmium (ppm)	2124	< 0.16		0.0342	0.0036	0.0028	8		518.43	5
518	Cadmium (ppm)	0689	< 0.2		0.0342	0.0036	0.0028	8		518.31	5
518	Cadmium (ppm)	0682	< 0.2		0.0342	0.0036	0.0028	8		518.43	5
520	Chromium (ppm)	0186	0.3044	0.0005	0.7986	0.3777	0.0975	10	-1.31	520.52	0
520	Chromium (ppm)	0968	0.4425	0.0110	0.7986	0.3777	0.0975	10	-0.94	520.43	0
520	Chromium (ppm)	2055	0.5210	0.1020	0.7986	0.3777	0.0975	10	-0.73	520.53	0
520	Chromium (ppm)	0164	0.5725	0.0130	0.7986	0.3777	0.0975	10	-0.60	520.53	0
520	Chromium (ppm)	0553	0.7170	0.2180	0.7986	0.3777	0.0975	10	-0.22	520.53	0
520	Chromium (ppm)	0693	0.9230	0.3320	0.7986	0.3777	0.0975	10	0.33	520.42	0
520	Chromium (ppm)	0918	1.001	0.0268	0.7986	0.3777	0.0975	10	0.54	520.53	0
520	Chromium (ppm)	0045	1.089	0.2020	0.7986	0.3777	0.0975	10	0.77	520.42	0
520	Chromium (ppm)	0510	1.115	0.0700	0.7986	0.3777	0.0975	10	0.84	520.43	0
520	Chromium (ppm)	0682	1.300	0.0000	0.7986	0.3777	0.0975	10	1.33	520.43	0
526	Lead (ppm)	0407	0.1138	0.0064	0.1851	0.0333	0.0081	7	-2.14	526.41	0
526	Lead (ppm)	0918	0.1749	0.0050	0.1851	0.0333	0.0081	7	-0.31	526.53	0
526	Lead (ppm)	0553	0.1775	0.0130	0.1851	0.0333	0.0081	7	-0.23	526.53	0
526	Lead (ppm)	0227	0.1830	0.0020	0.1851	0.0333	0.0081	7	-0.06	526.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
526	Lead (ppm)	0910	0.1850	0.0100	0.1851	0.0333	0.0081	7	0.00	526.52	0
526	Lead (ppm)	0407	0.2050	0.0040	0.1851	0.0333	0.0081	7	0.60	526.43	0
526	Lead (ppm)	2012	0.2350	0.0160	0.1851	0.0333	0.0081	7	1.50	526.53	0
526	Lead (ppm)	0689	< 0.2		0.1851	0.0333	0.0081	7		526.31	5
526	Lead (ppm)	0682	< 3		0.1851	0.0333	0.0081	7		526.43	5
529	Mercury (ppb)	2012	1.686	0.3870				2		529.99	0
529	Mercury (ppb)	0015	5,218	2,716				2		529.99	0
529	Mercury (ppb)	0227	< 10					2		529.99	5
529	Mercury (ppb)	0682	< 10000					2		529.99	5
539	Nickel (ppm)	0407	0.6863	0.0498	1.885	0.4822	0.1020	6	-2.49	539.41	0
539	Nickel (ppm)	0186	1.772	0.0077	1.885	0.4822	0.1020	6	-0.24	539.52	0
539	Nickel (ppm)	0682	1.900	0.0000	1.885	0.4822	0.1020	6	0.03	539.43	0
539	Nickel (ppm)	0553	1.950	0.2200	1.885	0.4822	0.1020	6	0.13	539.53	0
539	Nickel (ppm)	0918	2.068	0.0142	1.885	0.4822	0.1020	6	0.38	539.53	0
539	Nickel (ppm)	0015	2.460	0.3200	1.885	0.4822	0.1020	6	1.19	539.53	0
702	Butyric Acid (4:0) (%)	0619	0.0000	0.0000				0		702.00	4
704	Caproic Acid (6:0) (%)	0619	0.0000	0.0000				0		704.00	4
706	Caprylic acid (8:0) (%) (w/w)	0619	0.0000	0.0000				0		706.01	4
708	Capric acid (10:0) (%) (w/w)	0619	0.0160	0.0000				1		708.01	0
710	Lauric Acid (12:0) (%) (w/w)	0226	0.0075	0.0010				2		710.99	0
710	Lauric Acid (12:0) (%) (w/w)	0619	0.0215	0.0010				2		710.01	0
710	Lauric Acid (12:0) (%) (w/w)	2124	< 0.01					2		710.99	5
710	Lauric Acid (12:0) (%) (w/w)	0164	< 0.02					2		710.99	5
714	Myristic Acid (14:0) (%) (w/w)	0619	0.1165	0.0010	0.1484	0.0233	0.0008	4		714.01	0
714	Myristic Acid (14:0) (%) (w/w)	0297	0.1465	0.0010	0.1484	0.0233	0.0008	4		714.02	0
714	Myristic Acid (14:0) (%) (w/w)	0226	0.1605	0.0010	0.1484	0.0233	0.0008	4		714.99	0
714	Myristic Acid (14:0) (%) (w/w)	0164	0.1700	0.0000	0.1484	0.0233	0.0008	4		714.99	0
714	Myristic Acid (14:0) (%) (w/w)	2124	0.1450	0.0100	0.1484	0.0233	0.0008	4		714.99	1
716	Palmitic Acid (16:0) (%) (w/w)	2124	1.715	0.0100	2.005	0.2280	0.0100	4		716.99	0
716	Palmitic Acid (16:0) (%) (w/w)	0297	1.933	0.0100	2.005	0.2280	0.0100	4		716.02	0
716	Palmitic Acid (16:0) (%) (w/w)	0619	2.155	0.0100	2.005	0.2280	0.0100	4		716.01	0
716	Palmitic Acid (16:0) (%) (w/w)	0164	2.215	0.0100	2.005	0.2280	0.0100	4		716.99	0
716	Palmitic Acid (16:0) (%) (w/w)	0226	0.3150	0.0080	2.005	0.2280	0.0100	4		716.99	2
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0619	0.2400	0.0040	0.2671	0.0343	0.0118	4		718.01	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	2124	0.2450	0.0100	0.2671	0.0343	0.0118	4		718.99	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0297	0.2685	0.0250	0.2671	0.0343	0.0118	4		718.02	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0226	0.3150	0.0080	0.2671	0.0343	0.0118	4		718.99	0
720	Margaric acid (17:0) (%) (w/w)	0297	0.0729	0.0001				1		720.02	0
722	Stearic Acid (18:0) (%) (w/w)	2124	1.070	0.0200	1.149	0.0693	0.0100	5		722.99	0
722	Stearic Acid (18:0) (%) (w/w)	0297	1.087	0.0010	1.149	0.0693	0.0100	5		722.02	0
722	Stearic Acid (18:0) (%) (w/w)	0164	1.165	0.0100	1.149	0.0693	0.0100	5		722.99	0
722	Stearic Acid (18:0) (%) (w/w)	0619	1.190	0.0000	1.149	0.0693	0.0100	5		722.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
722	Stearic Acid (18:0) (% (w/w))	0226	1.234	0.0190	1.149	0.0693	0.0100	5		722.99	0
724	Oleic Acid (9c-18:1) (% (w/w))	2124	2.555	0.0500	3.220	0.6341	0.0462	5		724.99	0
724	Oleic Acid (9c-18:1) (% (w/w))	0297	2.806	0.0142	3.220	0.6341	0.0462	5		724.02	0
724	Oleic Acid (9c-18:1) (% (w/w))	0164	3.050	0.0600	3.220	0.6341	0.0462	5		724.99	0
724	Oleic Acid (9c-18:1) (% (w/w))	0226	3.538	0.0870	3.220	0.6341	0.0462	5		724.99	0
724	Oleic Acid (9c-18:1) (% (w/w))	0619	4.150	0.0200	3.220	0.6341	0.0462	5		724.01	0
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	2124	0.9750	0.0100	1.499	0.2180	0.0405	6	-2.40	726.99	0
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0297	1.376	0.0300	1.499	0.2180	0.0405	6	-0.56	726.02	0
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0148	1.525	0.0700	1.499	0.2180	0.0405	6	0.12	726.02	0
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0619	1.620	0.0000	1.499	0.2180	0.0405	6	0.56	726.01	0
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0164	1.620	0.0400	1.499	0.2180	0.0405	6	0.56	726.99	0
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0226	1.680	0.0930	1.499	0.2180	0.0405	6	0.83	726.99	0
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v)	2124	0.0500	0.0000	0.0773	0.0185	0.0044	5		728.99	0
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v)	0619	0.0695	0.0010	0.0773	0.0185	0.0044	5		728.01	0
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v)	0297	0.0790	0.0031	0.0773	0.0185	0.0044	5		728.02	0
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v)	0226	0.0930	0.0080	0.0773	0.0185	0.0044	5		728.99	0
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v)	0164	0.0950	0.0100	0.0773	0.0185	0.0044	5		728.99	0
730	Arachidic Acid (20:0) (% (w/w))	0226	0.0180	0.0020				2		730.99	0
730	Arachidic Acid (20:0) (% (w/w))	0619	0.0275	0.0010				2		730.01	0
730	Arachidic Acid (20:0) (% (w/w))	0297	0.0000	0.0000				2		730.02	4
730	Arachidic Acid (20:0) (% (w/w))	0164	< 0.02					2		730.99	5
732	Gondoic Acid (11c-20:1) (% (w/w))	0297	0.0235	0.0146	0.0286	0.0052	0.0087	4		732.02	0
732	Gondoic Acid (11c-20:1) (% (w/w))	0164	0.0250	0.0100	0.0286	0.0052	0.0087	4		732.99	0
732	Gondoic Acid (11c-20:1) (% (w/w))	0619	0.0320	0.0020	0.0286	0.0052	0.0087	4		732.01	0
732	Gondoic Acid (11c-20:1) (% (w/w))	0226	0.0340	0.0080	0.0286	0.0052	0.0087	4		732.99	0
736	Arachidonic Acid (5c,8c,11c,14c-20:4) (% (w/w)	0619	0.0415	0.0010				1		736.01	0
738	Mead Acid (11c,14c,17c-20:3) (% (w/w))	0619	0.0000	0.0000				0		738.01	4
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c,19c-20:5) (% (w/w))	0619	0.0000	0.0000				0		740.01	4
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c,19c-20:5) (% (w/w))	0226	< 0.005					0		740.99	5
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c,19c-20:5) (% (w/w))	2124	< 0.01					0		740.99	5
742	Behenic Acid (22:0) (% (w/w))	0226	0.0105	0.0010				1		742.99	0
742	Behenic Acid (22:0) (% (w/w))	0164	< 0.02					1		742.99	5
744	Erucic Acid (13c-22:1) (% (w/w))	0297	0.0484	0.0007				1		744.02	0
744	Erucic Acid (13c-22:1) (% (w/w))	0619	0.0000	0.0000				1		744.01	4
744	Erucic Acid (13c-22:1) (% (w/w))	0226	< 0.005					1		744.99	5
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c-22:5) (% (w/w))	0226	0.0050	0.0000				1		746.99	0
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c-22:5) (% (w/w))	0619	0.0000	0.0000				1		746.01	4
748	Lignoceric Acid (24:0) (% (w/w))	0226	0.0105	0.0030				1		748.99	0
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22c-22:6) (% (w/w))	0619	0.0000	0.0000				0		750.01	4
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22c-22:6) (% (w/w))	2171	0.0000	0.0000				0		750.02	4
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22c-22:6) (% (w/w))	0226	< 0.005					0		750.99	5
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22c-22:6) (% (w/w))	2124	< 0.01					0		750.99	5

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
752	Nervonic Acid (24:1) isomers (% (w/w))	0619	0.0000	0.0000				0	752.01	4	
752	Nervonic Acid (24:1) isomers (% (w/w))	0226	< 0.005					0	752.99	5	
754	Total n-3 Polyunsaturated (Omega-3) Fatty Aci	2124	0.0650	0.0100				3	754.99	0	
754	Total n-3 Polyunsaturated (Omega-3) Fatty Aci	0148	0.0989	0.0023				3	754.02	0	
754	Total n-3 Polyunsaturated (Omega-3) Fatty Aci	0164	0.1100	0.0000				3	754.99	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Aci	2124	1.040	0.0000	1.505	0.3144	0.0258	4	756.99	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Aci	0148	1.595	0.0700	1.505	0.3144	0.0258	4	756.01	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Aci	2171	1.659	0.0032	1.505	0.3144	0.0258	4	756.02	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Aci	0164	1.725	0.0300	1.505	0.3144	0.0258	4	756.99	0	
758	Total Saturated Fatty Acids (% (w/w))	0297	3.281	0.0160				3	758.02	0	
758	Total Saturated Fatty Acids (% (w/w))	2171	3.546	0.0150				3	758.02	0	
758	Total Saturated Fatty Acids (% (w/w))	0164	3.720	0.0200				3	758.99	0	
762	Total Monounsaturated Fatty Acids (% (w/w))	2171	3.375	0.0385				3	762.02	0	
762	Total Monounsaturated Fatty Acids (% (w/w))	0297	3.541	0.0056				3	762.02	0	
762	Total Monounsaturated Fatty Acids (% (w/w))	0164	3.710	0.0600				3	762.99	0	
766	Total Polyunsaturated Fatty Acids (% (w/w))	0297	1.487	0.0340				3	766.02	0	
766	Total Polyunsaturated Fatty Acids (% (w/w))	2171	1.758	0.0054				3	766.02	0	
766	Total Polyunsaturated Fatty Acids (% (w/w))	0164	1.860	0.0200				3	766.99	0	
770	Total Fat (equivalent to NLEA) (% (w/w))	0164	10.16	0.1100				1	770.99	0	
772	Total Fatty Acids (% (w/w))	2124	6.850	0.0400	8.573	1.106	0.1064	5	772.99	0	
772	Total Fatty Acids (% (w/w))	0297	8.310	0.0200	8.573	1.106	0.1064	5	772.02	0	
772	Total Fatty Acids (% (w/w))	2171	8.679	0.0590	8.573	1.106	0.1064	5	772.02	0	
772	Total Fatty Acids (% (w/w))	0226	9.322	0.3030	8.573	1.106	0.1064	5	772.99	0	
772	Total Fatty Acids (% (w/w))	0164	9.705	0.1100	8.573	1.106	0.1064	5	772.99	0	

Notes: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2 and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = Rejected for duplicates too far apart, 2 = Rejected as outlier, 5 = A reporting limit and 4 = zeros submitted as values. Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). Flag 3 indicates not used in statistics.