



AAFCO
Proficiency Testing Program



Animal Feed Scheme
Chicken Feed, Medicated
Test Material Code # 202029

Labs Reporting: 179
Analytes Reported 113
Issue Date : 10/31/2020

Analyte Proficiency Testing Report

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
000	Urea (%)	0278	0.4000	0.0000				1		000.02	0
001	Loss on Drying (%)	0853	6.750	0.5000	9.736	0.3559	0.1028	78	-8.39	001.07	0
001	Loss on Drying (%)	0019	7.050	0.3000	9.736	0.3559	0.1028	78	-7.55	001.07	0
001	Loss on Drying (%)	0618	8.020	0.0200	9.736	0.3559	0.1028	78	-4.82	001.99	0
001	Loss on Drying (%)	0028	8.390	0.0200	9.736	0.3559	0.1028	78	-3.78	001.07	0
001	Loss on Drying (%)	0536	8.870	0.0600	9.736	0.3559	0.1028	78	-2.43	001.99	0
001	Loss on Drying (%)	0357	8.955	0.0563	9.736	0.3559	0.1028	78	-2.19	001.99	0
001	Loss on Drying (%)	0889	8.965	0.0900	9.736	0.3559	0.1028	78	-2.17	001.99	0
001	Loss on Drying (%)	0345	9.100	0.0400	9.736	0.3559	0.1028	78	-1.79	001.07	0
001	Loss on Drying (%)	0278	9.155	0.0500	9.736	0.3559	0.1028	78	-1.63	001.07	0
001	Loss on Drying (%)	0034	9.165	0.0500	9.736	0.3559	0.1028	78	-1.60	001.07	0
001	Loss on Drying (%)	0027	9.257	0.0940	9.736	0.3559	0.1028	78	-1.34	001.00	0
001	Loss on Drying (%)	0918	9.270	0.0600	9.736	0.3559	0.1028	78	-1.31	001.99	0
001	Loss on Drying (%)	0878	9.370	0.0600	9.736	0.3559	0.1028	78	-1.03	001.03	0
001	Loss on Drying (%)	0027	9.394	0.0240	9.736	0.3559	0.1028	78	-0.96	001.07	0
001	Loss on Drying (%)	0610	9.400	0.0000	9.736	0.3559	0.1028	78	-0.94	001.99	0
001	Loss on Drying (%)	0089	9.420	0.0000	9.736	0.3559	0.1028	78	-0.89	001.07	0
001	Loss on Drying (%)	0581	9.420	0.0600	9.736	0.3559	0.1028	78	-0.89	001.07	0
001	Loss on Drying (%)	2144	9.505	0.1500	9.736	0.3559	0.1028	78	-0.65	001.07	0
001	Loss on Drying (%)	2246	9.525	0.0700	9.736	0.3559	0.1028	78	-0.59	001.99	0
001	Loss on Drying (%)	0226	9.540	0.0800	9.736	0.3559	0.1028	78	-0.55	001.07	0
001	Loss on Drying (%)	0693	9.566	0.5570	9.736	0.3559	0.1028	78	-0.48	001.07	0
001	Loss on Drying (%)	0948	9.595	0.0100	9.736	0.3559	0.1028	78	-0.40	001.99	0
001	Loss on Drying (%)	0682	9.600	0.0000	9.736	0.3559	0.1028	78	-0.38	001.07	0
001	Loss on Drying (%)	0590	9.620	0.2800	9.736	0.3559	0.1028	78	-0.32	001.07	0
001	Loss on Drying (%)	0021	9.650	0.5000	9.736	0.3559	0.1028	78	-0.24	001.99	0
001	Loss on Drying (%)	0142	9.655	0.0500	9.736	0.3559	0.1028	78	-0.23	001.07	0
001	Loss on Drying (%)	0656	9.660	0.0000	9.736	0.3559	0.1028	78	-0.21	001.99	0
001	Loss on Drying (%)	0510	9.700	0.2000	9.736	0.3559	0.1028	78	-0.10	001.99	0
001	Loss on Drying (%)	0066	9.710	0.0470	9.736	0.3559	0.1028	78	-0.07	001.07	0
001	Loss on Drying (%)	0353	9.710	0.0600	9.736	0.3559	0.1028	78	-0.07	001.07	0
001	Loss on Drying (%)	0169	9.715	0.0100	9.736	0.3559	0.1028	78	-0.06	001.00	0
001	Loss on Drying (%)	2109	9.715	0.0300	9.736	0.3559	0.1028	78	-0.06	001.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
001	Loss on Drying (%)	0037	9.725	0.0900	9.736	0.3559	0.1028	78	-0.03	001.99	0
001	Loss on Drying (%)	0560	9.725	0.5500	9.736	0.3559	0.1028	78	-0.03	001.99	0
001	Loss on Drying (%)	2058	9.730	0.0200	9.736	0.3559	0.1028	78	-0.02	001.03	0
001	Loss on Drying (%)	0683	9.735	0.0500	9.736	0.3559	0.1028	78	0.00	001.07	0
001	Loss on Drying (%)	2146	9.735	0.0500	9.736	0.3559	0.1028	78	0.00	001.07	0
001	Loss on Drying (%)	2059	9.740	0.0400	9.736	0.3559	0.1028	78	0.01	001.07	0
001	Loss on Drying (%)	0638	9.749	0.0780	9.736	0.3559	0.1028	78	0.04	001.07	0
001	Loss on Drying (%)	2319	9.750	0.1000	9.736	0.3559	0.1028	78	0.04	001.99	0
001	Loss on Drying (%)	0171	9.755	0.0100	9.736	0.3559	0.1028	78	0.05	001.07	0
001	Loss on Drying (%)	0074	9.765	0.1300	9.736	0.3559	0.1028	78	0.08	001.07	0
001	Loss on Drying (%)	0083	9.775	0.0100	9.736	0.3559	0.1028	78	0.11	001.07	0
001	Loss on Drying (%)	0366	9.800	0.2000	9.736	0.3559	0.1028	78	0.18	001.07	0
001	Loss on Drying (%)	0868	9.810	0.0400	9.736	0.3559	0.1028	78	0.21	001.03	0
001	Loss on Drying (%)	0571	9.810	0.0200	9.736	0.3559	0.1028	78	0.21	001.07	0
001	Loss on Drying (%)	0186	9.810	0.3200	9.736	0.3559	0.1028	78	0.21	001.99	0
001	Loss on Drying (%)	2009	9.821	0.0040	9.736	0.3559	0.1028	78	0.24	001.07	0
001	Loss on Drying (%)	0309	9.823	0.0070	9.736	0.3559	0.1028	78	0.24	001.00	0
001	Loss on Drying (%)	0505	9.830	0.1400	9.736	0.3559	0.1028	78	0.27	001.07	0
001	Loss on Drying (%)	2192	9.840	0.0400	9.736	0.3559	0.1028	78	0.29	001.07	0
001	Loss on Drying (%)	0662	9.846	0.0137	9.736	0.3559	0.1028	78	0.31	001.99	0
001	Loss on Drying (%)	0910	9.850	0.1000	9.736	0.3559	0.1028	78	0.32	001.99	0
001	Loss on Drying (%)	0872	9.860	0.0200	9.736	0.3559	0.1028	78	0.35	001.07	0
001	Loss on Drying (%)	0912	9.885	0.2100	9.736	0.3559	0.1028	78	0.42	001.99	0
001	Loss on Drying (%)	0619	9.895	0.0900	9.736	0.3559	0.1028	78	0.45	001.03	0
001	Loss on Drying (%)	2062	9.896	0.0108	9.736	0.3559	0.1028	78	0.45	001.03	0
001	Loss on Drying (%)	0297	9.910	0.2200	9.736	0.3559	0.1028	78	0.49	001.07	0
001	Loss on Drying (%)	0650	9.960	0.0600	9.736	0.3559	0.1028	78	0.63	001.07	0
001	Loss on Drying (%)	0035	9.965	0.0300	9.736	0.3559	0.1028	78	0.64	001.07	0
001	Loss on Drying (%)	0019	9.985	0.0700	9.736	0.3559	0.1028	78	0.70	001.99	0
001	Loss on Drying (%)	0590	10.00	0.0000	9.736	0.3559	0.1028	78	0.74	001.99	0
001	Loss on Drying (%)	0644	10.04	0.0100	9.736	0.3559	0.1028	78	0.84	001.05	0
001	Loss on Drying (%)	0407	10.04	0.0035	9.736	0.3559	0.1028	78	0.85	001.07	0
001	Loss on Drying (%)	0876	10.05	0.1000	9.736	0.3559	0.1028	78	0.88	001.07	0
001	Loss on Drying (%)	2295	10.09	0.0100	9.736	0.3559	0.1028	78	0.98	001.03	0
001	Loss on Drying (%)	0003	10.09	0.0100	9.736	0.3559	0.1028	78	0.98	001.07	0
001	Loss on Drying (%)	0148	10.09	0.2200	9.736	0.3559	0.1028	78	1.00	001.07	0
001	Loss on Drying (%)	0689	10.10	0.2000	9.736	0.3559	0.1028	78	1.02	001.07	0
001	Loss on Drying (%)	0098	10.11	0.1000	9.736	0.3559	0.1028	78	1.05	001.07	0
001	Loss on Drying (%)	0049	10.13	0.1900	9.736	0.3559	0.1028	78	1.09	001.07	0
001	Loss on Drying (%)	0183	10.25	0.0310	9.736	0.3559	0.1028	78	1.44	001.00	0
001	Loss on Drying (%)	0045	10.31	0.0600	9.736	0.3559	0.1028	78	1.61	001.07	0
001	Loss on Drying (%)	0893	10.33	0.0300	9.736	0.3559	0.1028	78	1.66	001.03	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
001	Loss on Drying (%)	0033	10.34	0.3100	9.736	0.3559	0.1028	78	1.68	001.07	0
001	Loss on Drying (%)	0843	10.41	0.1600	9.736	0.3559	0.1028	78	1.90	001.03	0
001	Loss on Drying (%)	0843	10.41	0.1600	9.736	0.3559	0.1028	78	1.90	001.07	0
001	Loss on Drying (%)	2328	10.56	0.1700	9.736	0.3559	0.1028	78	2.30	001.99	0
001	Loss on Drying (%)	2155	8.995	0.6900	9.736	0.3559	0.1028	78	-2.08	001.99	1
001	Loss on Drying (%)	2081	9.400	0.6600	9.736	0.3559	0.1028	78	-0.94	001.99	1
002	Protein, Crude (%)	0910	18.85	0.5000	19.88	0.2513	0.1549	170	-4.11	002.01	0
002	Protein, Crude (%)	0852	19.20	0.0300	19.88	0.2513	0.1549	170	-2.74	002.11	0
002	Protein, Crude (%)	0953	19.24	0.0100	19.88	0.2513	0.1549	170	-2.58	002.06	0
002	Protein, Crude (%)	2291	19.32	0.1900	19.88	0.2513	0.1549	170	-2.26	002.05	0
002	Protein, Crude (%)	2109	19.35	0.0700	19.88	0.2513	0.1549	170	-2.14	002.05	0
002	Protein, Crude (%)	0148	19.40	0.2000	19.88	0.2513	0.1549	170	-1.92	002.11	0
002	Protein, Crude (%)	0656	19.41	0.0000	19.88	0.2513	0.1549	170	-1.88	002.01	0
002	Protein, Crude (%)	0089	19.47	0.0100	19.88	0.2513	0.1549	170	-1.66	002.06	0
002	Protein, Crude (%)	2328	19.50	0.0100	19.88	0.2513	0.1549	170	-1.51	002.00	0
002	Protein, Crude (%)	0889	19.51	0.0800	19.88	0.2513	0.1549	170	-1.49	002.11	0
002	Protein, Crude (%)	0407	19.51	0.5060	19.88	0.2513	0.1549	170	-1.49	002.06	0
002	Protein, Crude (%)	0036	19.53	0.0374	19.88	0.2513	0.1549	170	-1.39	002.02	0
002	Protein, Crude (%)	0563	19.53	0.0765	19.88	0.2513	0.1549	170	-1.39	002.08	0
002	Protein, Crude (%)	0968	19.54	0.0500	19.88	0.2513	0.1549	170	-1.39	002.01	0
002	Protein, Crude (%)	0553	19.54	0.0600	19.88	0.2513	0.1549	170	-1.37	002.11	0
002	Protein, Crude (%)	2295	19.55	0.0100	19.88	0.2513	0.1549	170	-1.35	002.05	0
002	Protein, Crude (%)	2062	19.55	0.0579	19.88	0.2513	0.1549	170	-1.32	002.05	0
002	Protein, Crude (%)	0074	19.56	0.4800	19.88	0.2513	0.1549	170	-1.29	002.06	0
002	Protein, Crude (%)	0019	19.57	0.0400	19.88	0.2513	0.1549	170	-1.25	002.06	0
002	Protein, Crude (%)	2058	19.58	0.0600	19.88	0.2513	0.1549	170	-1.21	002.05	0
002	Protein, Crude (%)	0183	19.59	0.0800	19.88	0.2513	0.1549	170	-1.17	002.01	0
002	Protein, Crude (%)	0098	19.59	0.1200	19.88	0.2513	0.1549	170	-1.17	002.08	0
002	Protein, Crude (%)	0689	19.60	0.2000	19.88	0.2513	0.1549	170	-1.13	002.05	0
002	Protein, Crude (%)	0033	19.60	0.0000	19.88	0.2513	0.1549	170	-1.13	002.06	0
002	Protein, Crude (%)	2181	19.61	0.3000	19.88	0.2513	0.1549	170	-1.09	002.05	0
002	Protein, Crude (%)	0171	19.62	0.1000	19.88	0.2513	0.1549	170	-1.05	002.06	0
002	Protein, Crude (%)	2302	19.63	0.1300	19.88	0.2513	0.1549	170	-1.03	002.06	0
002	Protein, Crude (%)	0018	19.64	0.4900	19.88	0.2513	0.1549	170	-0.99	002.06	0
002	Protein, Crude (%)	0539	19.64	0.2300	19.88	0.2513	0.1549	170	-0.99	002.06	0
002	Protein, Crude (%)	0066	19.64	0.0730	19.88	0.2513	0.1549	170	-0.96	002.02	0
002	Protein, Crude (%)	0202	19.65	0.0700	19.88	0.2513	0.1549	170	-0.95	002.06	0
002	Protein, Crude (%)	0870	19.65	0.2400	19.88	0.2513	0.1549	170	-0.93	002.01	0
002	Protein, Crude (%)	0893	19.69	0.0600	19.88	0.2513	0.1549	170	-0.77	002.05	0
002	Protein, Crude (%)	0598	19.69	0.0200	19.88	0.2513	0.1549	170	-0.77	002.06	0
002	Protein, Crude (%)	2193	19.69	0.0400	19.88	0.2513	0.1549	170	-0.77	002.01	0
002	Protein, Crude (%)	0425	19.69	0.0000	19.88	0.2513	0.1549	170	-0.77	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0683	19.70	0.2700	19.88	0.2513	0.1549	170	-0.75	002.05	0
002	Protein, Crude (%)	0652	19.70	0.0000	19.88	0.2513	0.1549	170	-0.73	002.01	0
002	Protein, Crude (%)	0910	19.70	0.0000	19.88	0.2513	0.1549	170	-0.73	002.06	0
002	Protein, Crude (%)	0529	19.70	0.1200	19.88	0.2513	0.1549	170	-0.73	002.06	0
002	Protein, Crude (%)	0897	19.71	0.0100	19.88	0.2513	0.1549	170	-0.71	002.05	0
002	Protein, Crude (%)	2006	19.72	0.1200	19.88	0.2513	0.1549	170	-0.65	002.05	0
002	Protein, Crude (%)	2168	19.72	0.0400	19.88	0.2513	0.1549	170	-0.65	002.06	0
002	Protein, Crude (%)	2073	19.74	0.2600	19.88	0.2513	0.1549	170	-0.57	002.05	0
002	Protein, Crude (%)	0885	19.75	0.0675	19.88	0.2513	0.1549	170	-0.55	002.05	0
002	Protein, Crude (%)	0164	19.75	0.1200	19.88	0.2513	0.1549	170	-0.53	002.01	0
002	Protein, Crude (%)	2146	19.75	0.1200	19.88	0.2513	0.1549	170	-0.53	002.05	0
002	Protein, Crude (%)	0229	19.75	0.5800	19.88	0.2513	0.1549	170	-0.53	002.06	0
002	Protein, Crude (%)	0692	19.75	0.7000	19.88	0.2513	0.1549	170	-0.53	002.06	0
002	Protein, Crude (%)	0884	19.75	0.1000	19.88	0.2513	0.1549	170	-0.53	002.06	0
002	Protein, Crude (%)	0728	19.75	0.0040	19.88	0.2513	0.1549	170	-0.52	002.04	0
002	Protein, Crude (%)	2089	19.76	0.1100	19.88	0.2513	0.1549	170	-0.51	002.06	0
002	Protein, Crude (%)	0242	19.76	0.0100	19.88	0.2513	0.1549	170	-0.51	002.06	0
002	Protein, Crude (%)	0001	19.76	0.2400	19.88	0.2513	0.1549	170	-0.49	002.06	0
002	Protein, Crude (%)	0868	19.77	0.0800	19.88	0.2513	0.1549	170	-0.45	002.05	0
002	Protein, Crude (%)	2129	19.77	0.1700	19.88	0.2513	0.1549	170	-0.45	002.06	0
002	Protein, Crude (%)	0948	19.78	0.0200	19.88	0.2513	0.1549	170	-0.41	002.05	0
002	Protein, Crude (%)	0049	19.78	0.5600	19.88	0.2513	0.1549	170	-0.41	002.06	0
002	Protein, Crude (%)	0263	19.80	0.0070	19.88	0.2513	0.1549	170	-0.35	002.06	0
002	Protein, Crude (%)	2009	19.80	0.0200	19.88	0.2513	0.1549	170	-0.34	002.05	0
002	Protein, Crude (%)	0619	19.80	0.2000	19.88	0.2513	0.1549	170	-0.33	002.05	0
002	Protein, Crude (%)	0619	19.80	0.2000	19.88	0.2513	0.1549	170	-0.33	002.06	0
002	Protein, Crude (%)	0968	19.80	0.1200	19.88	0.2513	0.1549	170	-0.33	002.06	0
002	Protein, Crude (%)	2199	19.80	0.4000	19.88	0.2513	0.1549	170	-0.33	002.05	0
002	Protein, Crude (%)	0123	19.80	0.1000	19.88	0.2513	0.1549	170	-0.33	002.06	0
002	Protein, Crude (%)	0881	19.80	0.0120	19.88	0.2513	0.1549	170	-0.32	002.05	0
002	Protein, Crude (%)	2022	19.81	0.1100	19.88	0.2513	0.1549	170	-0.31	002.05	0
002	Protein, Crude (%)	0353	19.81	0.1900	19.88	0.2513	0.1549	170	-0.31	002.06	0
002	Protein, Crude (%)	0505	19.81	0.0100	19.88	0.2513	0.1549	170	-0.31	002.06	0
002	Protein, Crude (%)	0226	19.81	0.1800	19.88	0.2513	0.1549	170	-0.29	002.06	0
002	Protein, Crude (%)	2192	19.81	0.0600	19.88	0.2513	0.1549	170	-0.29	002.05	0
002	Protein, Crude (%)	0028	19.82	0.0200	19.88	0.2513	0.1549	170	-0.25	002.06	0
002	Protein, Crude (%)	0144	19.83	0.0400	19.88	0.2513	0.1549	170	-0.21	002.06	0
002	Protein, Crude (%)	0939	19.84	0.2600	19.88	0.2513	0.1549	170	-0.17	002.05	0
002	Protein, Crude (%)	0650	19.84	0.0200	19.88	0.2513	0.1549	170	-0.17	002.06	0
002	Protein, Crude (%)	2196	19.85	0.0000	19.88	0.2513	0.1549	170	-0.13	002.01	0
002	Protein, Crude (%)	0687	19.85	0.1000	19.88	0.2513	0.1549	170	-0.13	002.06	0
002	Protein, Crude (%)	0876	19.85	0.1000	19.88	0.2513	0.1549	170	-0.13	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0859	19.86	0.2980	19.88	0.2513	0.1549	170	-0.08	002.06	0
002	Protein, Crude (%)	2188	19.88	0.0900	19.88	0.2513	0.1549	170	-0.03	002.01	0
002	Protein, Crude (%)	0164	19.88	0.1300	19.88	0.2513	0.1549	170	-0.03	002.06	0
002	Protein, Crude (%)	0357	19.88	0.1250	19.88	0.2513	0.1549	170	-0.03	002.06	0
002	Protein, Crude (%)	0870	19.88	0.0294	19.88	0.2513	0.1549	170	-0.02	002.06	0
002	Protein, Crude (%)	0036	19.88	0.1400	19.88	0.2513	0.1549	170	-0.01	002.06	0
002	Protein, Crude (%)	0297	19.88	0.0800	19.88	0.2513	0.1549	170	-0.01	002.06	0
002	Protein, Crude (%)	2081	19.89	0.1600	19.88	0.2513	0.1549	170	0.03	002.06	0
002	Protein, Crude (%)	2200	19.90	0.0000	19.88	0.2513	0.1549	170	0.07	002.05	0
002	Protein, Crude (%)	0010	19.90	0.2000	19.88	0.2513	0.1549	170	0.07	002.06	0
002	Protein, Crude (%)	0278	19.90	0.0000	19.88	0.2513	0.1549	170	0.07	002.06	0
002	Protein, Crude (%)	0298	19.90	0.0200	19.88	0.2513	0.1549	170	0.07	002.06	0
002	Protein, Crude (%)	0345	19.90	0.6000	19.88	0.2513	0.1549	170	0.07	002.06	0
002	Protein, Crude (%)	0590	19.90	0.0000	19.88	0.2513	0.1549	170	0.07	002.06	0
002	Protein, Crude (%)	0853	19.90	0.2000	19.88	0.2513	0.1549	170	0.07	002.06	0
002	Protein, Crude (%)	0083	19.91	0.1300	19.88	0.2513	0.1549	170	0.09	002.06	0
002	Protein, Crude (%)	0589	19.91	0.0900	19.88	0.2513	0.1549	170	0.09	002.06	0
002	Protein, Crude (%)	0878	19.92	0.0000	19.88	0.2513	0.1549	170	0.15	002.05	0
002	Protein, Crude (%)	0027	19.92	0.1230	19.88	0.2513	0.1549	170	0.16	002.06	0
002	Protein, Crude (%)	0682	19.93	0.0000	19.88	0.2513	0.1549	170	0.19	002.06	0
002	Protein, Crude (%)	0939	19.94	0.0700	19.88	0.2513	0.1549	170	0.20	002.06	0
002	Protein, Crude (%)	0035	19.94	0.0300	19.88	0.2513	0.1549	170	0.20	002.06	0
002	Protein, Crude (%)	2216	19.94	0.0900	19.88	0.2513	0.1549	170	0.20	002.06	0
002	Protein, Crude (%)	0693	19.94	0.3750	19.88	0.2513	0.1549	170	0.21	002.06	0
002	Protein, Crude (%)	0148	19.95	0.1000	19.88	0.2513	0.1549	170	0.26	002.06	0
002	Protein, Crude (%)	0610	19.95	0.1000	19.88	0.2513	0.1549	170	0.26	002.06	0
002	Protein, Crude (%)	0100	19.95	0.4400	19.88	0.2513	0.1549	170	0.26	002.06	0
002	Protein, Crude (%)	0553	19.95	0.1600	19.88	0.2513	0.1549	170	0.26	002.06	0
002	Protein, Crude (%)	2150	19.95	0.0600	19.88	0.2513	0.1549	170	0.26	002.06	0
002	Protein, Crude (%)	0034	19.96	0.2660	19.88	0.2513	0.1549	170	0.29	002.06	0
002	Protein, Crude (%)	2144	19.96	0.2600	19.88	0.2513	0.1549	170	0.30	002.05	0
002	Protein, Crude (%)	0662	19.97	0.3125	19.88	0.2513	0.1549	170	0.34	002.01	0
002	Protein, Crude (%)	0358	19.98	0.3100	19.88	0.2513	0.1549	170	0.36	002.06	0
002	Protein, Crude (%)	0868	19.98	0.0100	19.88	0.2513	0.1549	170	0.36	002.06	0
002	Protein, Crude (%)	0644	19.98	0.2010	19.88	0.2513	0.1549	170	0.39	002.06	0
002	Protein, Crude (%)	0309	19.99	0.4540	19.88	0.2513	0.1549	170	0.42	002.06	0
002	Protein, Crude (%)	0948	19.99	0.0400	19.88	0.2513	0.1549	170	0.42	002.06	0
002	Protein, Crude (%)	0964	20.00	0.1760	19.88	0.2513	0.1549	170	0.45	002.06	0
002	Protein, Crude (%)	0638	20.00	0.0000	19.88	0.2513	0.1549	170	0.46	002.04	0
002	Protein, Crude (%)	0520	20.00	0.2000	19.88	0.2513	0.1549	170	0.46	002.06	0
002	Protein, Crude (%)	0914	20.01	0.0500	19.88	0.2513	0.1549	170	0.48	002.06	0
002	Protein, Crude (%)	0626	20.01	0.0050	19.88	0.2513	0.1549	170	0.49	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0006	20.01	0.2800	19.88	0.2513	0.1549	170	0.50	002.06	0
002	Protein, Crude (%)	0918	20.01	0.5600	19.88	0.2513	0.1549	170	0.50	002.06	0
002	Protein, Crude (%)	2174	20.02	0.1000	19.88	0.2513	0.1549	170	0.54	002.06	0
002	Protein, Crude (%)	0009	20.04	0.2100	19.88	0.2513	0.1549	170	0.60	002.06	0
002	Protein, Crude (%)	0142	20.04	0.1500	19.88	0.2513	0.1549	170	0.60	002.06	0
002	Protein, Crude (%)	0186	20.05	0.5300	19.88	0.2513	0.1549	170	0.64	002.06	0
002	Protein, Crude (%)	0571	20.05	0.1700	19.88	0.2513	0.1549	170	0.64	002.06	0
002	Protein, Crude (%)	0045	20.05	0.3000	19.88	0.2513	0.1549	170	0.66	002.06	0
002	Protein, Crude (%)	0175	20.05	0.1000	19.88	0.2513	0.1549	170	0.66	002.06	0
002	Protein, Crude (%)	2246	20.07	0.3100	19.88	0.2513	0.1549	170	0.72	002.06	0
002	Protein, Crude (%)	0508	20.08	0.2400	19.88	0.2513	0.1549	170	0.77	002.06	0
002	Protein, Crude (%)	0108	20.09	0.0600	19.88	0.2513	0.1549	170	0.82	002.06	0
002	Protein, Crude (%)	2319	20.09	0.0200	19.88	0.2513	0.1549	170	0.82	002.06	0
002	Protein, Crude (%)	0208	20.10	0.2000	19.88	0.2513	0.1549	170	0.86	002.06	0
002	Protein, Crude (%)	0265	20.10	0.6000	19.88	0.2513	0.1549	170	0.86	002.06	0
002	Protein, Crude (%)	0574	20.10	0.3600	19.88	0.2513	0.1549	170	0.86	002.06	0
002	Protein, Crude (%)	0590	20.10	0.0000	19.88	0.2513	0.1549	170	0.86	002.11	0
002	Protein, Crude (%)	0941	20.11	0.2500	19.88	0.2513	0.1549	170	0.88	002.05	0
002	Protein, Crude (%)	0037	20.11	0.1000	19.88	0.2513	0.1549	170	0.90	002.06	0
002	Protein, Crude (%)	0541	20.11	0.1400	19.88	0.2513	0.1549	170	0.90	002.06	0
002	Protein, Crude (%)	2290	20.11	0.2800	19.88	0.2513	0.1549	170	0.90	002.11	0
002	Protein, Crude (%)	0004	20.12	0.4300	19.88	0.2513	0.1549	170	0.92	002.06	0
002	Protein, Crude (%)	2213	20.12	0.0200	19.88	0.2513	0.1549	170	0.94	002.06	0
002	Protein, Crude (%)	0003	20.13	0.1900	19.88	0.2513	0.1549	170	0.96	002.06	0
002	Protein, Crude (%)	0872	20.13	0.1810	19.88	0.2513	0.1549	170	0.97	002.06	0
002	Protein, Crude (%)	0921	20.13	0.0320	19.88	0.2513	0.1549	170	0.98	002.06	0
002	Protein, Crude (%)	2076	20.14	0.1530	19.88	0.2513	0.1549	170	1.04	002.06	0
002	Protein, Crude (%)	0190	20.15	0.0900	19.88	0.2513	0.1549	170	1.04	002.06	0
002	Protein, Crude (%)	0294	20.15	0.1000	19.88	0.2513	0.1549	170	1.06	002.06	0
002	Protein, Crude (%)	0510	20.15	0.1000	19.88	0.2513	0.1549	170	1.06	002.06	0
002	Protein, Crude (%)	0536	20.17	0.0300	19.88	0.2513	0.1549	170	1.12	002.05	0
002	Protein, Crude (%)	0656	20.19	0.0000	19.88	0.2513	0.1549	170	1.22	002.06	0
002	Protein, Crude (%)	2303	20.22	0.1400	19.88	0.2513	0.1549	170	1.34	002.06	0
002	Protein, Crude (%)	0098	20.25	0.3000	19.88	0.2513	0.1549	170	1.46	002.06	0
002	Protein, Crude (%)	0726	20.28	0.0500	19.88	0.2513	0.1549	170	1.56	002.06	0
002	Protein, Crude (%)	2326	20.28	0.0600	19.88	0.2513	0.1549	170	1.58	002.06	0
002	Protein, Crude (%)	0889	20.29	0.0900	19.88	0.2513	0.1549	170	1.60	002.01	0
002	Protein, Crude (%)	0047	20.35	0.5000	19.88	0.2513	0.1549	170	1.86	002.06	0
002	Protein, Crude (%)	0169	20.39	0.1800	19.88	0.2513	0.1549	170	2.02	002.00	0
002	Protein, Crude (%)	0042	20.40	0.4000	19.88	0.2513	0.1549	170	2.06	002.06	0
002	Protein, Crude (%)	2059	20.42	0.0407	19.88	0.2513	0.1549	170	2.12	002.06	0
002	Protein, Crude (%)	0366	20.45	0.5000	19.88	0.2513	0.1549	170	2.25	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0852	20.49	0.1800	19.88	0.2513	0.1549	170	2.41	002.05	0
002	Protein, Crude (%)	0297	20.53	0.1300	19.88	0.2513	0.1549	170	2.55	002.11	0
002	Protein, Crude (%)	2155	20.61	0.1450	19.88	0.2513	0.1549	170	2.90	002.06	0
002	Protein, Crude (%)	0685	20.69	0.0100	19.88	0.2513	0.1549	170	3.19	002.01	0
002	Protein, Crude (%)	0618	20.73	0.1100	19.88	0.2513	0.1549	170	3.35	002.06	0
002	Protein, Crude (%)	0581	20.75	0.2100	19.88	0.2513	0.1549	170	3.43	002.05	0
002	Protein, Crude (%)	0011	20.80	0.4000	19.88	0.2513	0.1549	170	3.65	002.11	0
002	Protein, Crude (%)	2302	20.97	0.0000	19.88	0.2513	0.1549	170	4.32	002.11	0
002	Protein, Crude (%)	2184	18.00	0.9600	19.88	0.2513	0.1549	170	-7.49	002.06	1
002	Protein, Crude (%)	2329	18.99	1.080	19.88	0.2513	0.1549	170	-3.55	002.01	1
002	Protein, Crude (%)	2178	19.90	1.050	19.88	0.2513	0.1549	170	0.05	002.06	1
002	Protein, Crude (%)	0843	20.28	1.040	19.88	0.2513	0.1549	170	1.58	002.06	1
002	Protein, Crude (%)	0643	16.28	0.1200	19.88	0.2513	0.1549	170	-14.34	002.99	2
002	Protein, Crude (%)	0536	16.53	0.1200	19.88	0.2513	0.1549	170	-13.34	002.03	2
003	Fat, Crude (%)	2291	1.515	0.0900	2.469	0.2109	0.0928	143	-4.52	003.13	0
003	Fat, Crude (%)	0889	1.700	0.0800	2.469	0.2109	0.0928	143	-3.65	003.11	0
003	Fat, Crude (%)	0265	1.715	0.3700	2.469	0.2109	0.0928	143	-3.57	003.14	0
003	Fat, Crude (%)	2246	1.730	0.0000	2.469	0.2109	0.0928	143	-3.50	003.13	0
003	Fat, Crude (%)	0108	1.740	0.1000	2.469	0.2109	0.0928	143	-3.46	003.14	0
003	Fat, Crude (%)	0175	1.880	0.0400	2.469	0.2109	0.0928	143	-2.79	003.14	0
003	Fat, Crude (%)	2146	2.030	0.2400	2.469	0.2109	0.0928	143	-2.08	003.14	0
003	Fat, Crude (%)	2089	2.050	0.0400	2.469	0.2109	0.0928	143	-1.99	003.10	0
003	Fat, Crude (%)	2143	2.080	0.0200	2.469	0.2109	0.0928	143	-1.84	003.14	0
003	Fat, Crude (%)	2073	2.105	0.0500	2.469	0.2109	0.0928	143	-1.73	003.10	0
003	Fat, Crude (%)	0183	2.108	0.1760	2.469	0.2109	0.0928	143	-1.71	003.14	0
003	Fat, Crude (%)	2144	2.130	0.1000	2.469	0.2109	0.0928	143	-1.61	003.14	0
003	Fat, Crude (%)	0144	2.155	0.4300	2.469	0.2109	0.0928	143	-1.49	003.14	0
003	Fat, Crude (%)	0037	2.170	0.0600	2.469	0.2109	0.0928	143	-1.42	003.14	0
003	Fat, Crude (%)	0693	2.173	0.1300	2.469	0.2109	0.0928	143	-1.40	003.10	0
003	Fat, Crude (%)	0510	2.180	0.0000	2.469	0.2109	0.0928	143	-1.37	003.14	0
003	Fat, Crude (%)	0581	2.225	0.1100	2.469	0.2109	0.0928	143	-1.16	003.14	0
003	Fat, Crude (%)	0034	2.225	0.0500	2.469	0.2109	0.0928	143	-1.16	003.14	0
003	Fat, Crude (%)	2302	2.230	0.2200	2.469	0.2109	0.0928	143	-1.13	003.14	0
003	Fat, Crude (%)	0042	2.245	0.1900	2.469	0.2109	0.0928	143	-1.06	003.10	0
003	Fat, Crude (%)	0186	2.250	0.1000	2.469	0.2109	0.0928	143	-1.04	003.14	0
003	Fat, Crude (%)	0876	2.250	0.1000	2.469	0.2109	0.0928	143	-1.04	003.14	0
003	Fat, Crude (%)	0536	2.265	0.1300	2.469	0.2109	0.0928	143	-0.97	003.99	0
003	Fat, Crude (%)	2193	2.270	0.0200	2.469	0.2109	0.0928	143	-0.94	003.10	0
003	Fat, Crude (%)	0036	2.289	0.1388	2.469	0.2109	0.0928	143	-0.85	003.10	0
003	Fat, Crude (%)	2022	2.290	0.0400	2.469	0.2109	0.0928	143	-0.85	003.10	0
003	Fat, Crude (%)	2009	2.291	0.0135	2.469	0.2109	0.0928	143	-0.84	003.10	0
003	Fat, Crude (%)	0357	2.300	0.0000	2.469	0.2109	0.0928	143	-0.80	003.14	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
003	Fat, Crude (%)	0529	2.310	0.1200	2.469	0.2109	0.0928	143	-0.75	003.14	0
003	Fat, Crude (%)	0939	2.315	0.1700	2.469	0.2109	0.0928	143	-0.73	003.06	0
003	Fat, Crude (%)	0019	2.325	0.0300	2.469	0.2109	0.0928	143	-0.68	003.14	0
003	Fat, Crude (%)	0148	2.330	0.0600	2.469	0.2109	0.0928	143	-0.66	003.14	0
003	Fat, Crude (%)	2184	2.335	0.0500	2.469	0.2109	0.0928	143	-0.63	003.10	0
003	Fat, Crude (%)	2326	2.335	0.0500	2.469	0.2109	0.0928	143	-0.63	003.12	0
003	Fat, Crude (%)	0912	2.335	0.0900	2.469	0.2109	0.0928	143	-0.63	003.14	0
003	Fat, Crude (%)	0263	2.350	0.0370	2.469	0.2109	0.0928	143	-0.57	003.10	0
003	Fat, Crude (%)	0505	2.350	0.1000	2.469	0.2109	0.0928	143	-0.56	003.14	0
003	Fat, Crude (%)	0689	2.350	0.3000	2.469	0.2109	0.0928	143	-0.56	003.10	0
003	Fat, Crude (%)	0868	2.350	0.0400	2.469	0.2109	0.0928	143	-0.56	003.14	0
003	Fat, Crude (%)	0033	2.355	0.0100	2.469	0.2109	0.0928	143	-0.54	003.13	0
003	Fat, Crude (%)	0358	2.355	0.1700	2.469	0.2109	0.0928	143	-0.54	003.14	0
003	Fat, Crude (%)	0726	2.355	0.0500	2.469	0.2109	0.0928	143	-0.54	003.14	0
003	Fat, Crude (%)	0897	2.375	0.0100	2.469	0.2109	0.0928	143	-0.44	003.10	0
003	Fat, Crude (%)	0948	2.380	0.0600	2.469	0.2109	0.0928	143	-0.42	003.14	0
003	Fat, Crude (%)	0035	2.390	0.0600	2.469	0.2109	0.0928	143	-0.37	003.14	0
003	Fat, Crude (%)	0242	2.395	0.0500	2.469	0.2109	0.0928	143	-0.35	003.10	0
003	Fat, Crude (%)	0683	2.395	0.1100	2.469	0.2109	0.0928	143	-0.35	003.14	0
003	Fat, Crude (%)	0918	2.400	0.0400	2.469	0.2109	0.0928	143	-0.33	003.06	0
003	Fat, Crude (%)	0728	2.400	0.2000	2.469	0.2109	0.0928	143	-0.33	003.10	0
003	Fat, Crude (%)	2196	2.400	0.0000	2.469	0.2109	0.0928	143	-0.33	003.10	0
003	Fat, Crude (%)	0598	2.400	0.0200	2.469	0.2109	0.0928	143	-0.33	003.14	0
003	Fat, Crude (%)	0171	2.405	0.0900	2.469	0.2109	0.0928	143	-0.30	003.14	0
003	Fat, Crude (%)	0878	2.405	0.0300	2.469	0.2109	0.0928	143	-0.30	003.14	0
003	Fat, Crude (%)	2129	2.407	0.2140	2.469	0.2109	0.0928	143	-0.29	003.14	0
003	Fat, Crude (%)	0407	2.429	0.1025	2.469	0.2109	0.0928	143	-0.19	003.14	0
003	Fat, Crude (%)	2155	2.433	0.1830	2.469	0.2109	0.0928	143	-0.17	003.10	0
003	Fat, Crude (%)	0910	2.435	0.0900	2.469	0.2109	0.0928	143	-0.16	003.06	0
003	Fat, Crude (%)	0948	2.435	0.0100	2.469	0.2109	0.0928	143	-0.16	003.09	0
003	Fat, Crude (%)	0045	2.435	0.0300	2.469	0.2109	0.0928	143	-0.16	003.10	0
003	Fat, Crude (%)	0619	2.440	0.1600	2.469	0.2109	0.0928	143	-0.14	003.10	0
003	Fat, Crude (%)	0001	2.441	0.0489	2.469	0.2109	0.0928	143	-0.13	003.14	0
003	Fat, Crude (%)	0638	2.441	0.0660	2.469	0.2109	0.0928	143	-0.13	003.09	0
003	Fat, Crude (%)	2006	2.445	0.0300	2.469	0.2109	0.0928	143	-0.11	003.13	0
003	Fat, Crude (%)	0229	2.445	0.0500	2.469	0.2109	0.0928	143	-0.11	003.14	0
003	Fat, Crude (%)	0366	2.450	0.1000	2.469	0.2109	0.0928	143	-0.09	003.10	0
003	Fat, Crude (%)	0028	2.450	0.0000	2.469	0.2109	0.0928	143	-0.09	003.13	0
003	Fat, Crude (%)	0098	2.450	0.0400	2.469	0.2109	0.0928	143	-0.09	003.13	0
003	Fat, Crude (%)	0692	2.450	0.1000	2.469	0.2109	0.0928	143	-0.09	003.14	0
003	Fat, Crude (%)	0098	2.455	0.0700	2.469	0.2109	0.0928	143	-0.07	003.10	0
003	Fat, Crude (%)	0123	2.455	0.0500	2.469	0.2109	0.0928	143	-0.07	003.13	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
003	Fat, Crude (%)	0870	2.459	0.0772	2.469	0.2109	0.0928	143	-0.05	003.10	0
003	Fat, Crude (%)	0541	2.460	0.1200	2.469	0.2109	0.0928	143	-0.04	003.10	0
003	Fat, Crude (%)	0425	2.465	0.0500	2.469	0.2109	0.0928	143	-0.02	003.06	0
003	Fat, Crude (%)	2181	2.470	0.0600	2.469	0.2109	0.0928	143	0.01	003.06	0
003	Fat, Crude (%)	0098	2.470	0.0400	2.469	0.2109	0.0928	143	0.01	003.09	0
003	Fat, Crude (%)	0563	2.474	0.2273	2.469	0.2109	0.0928	143	0.03	003.06	0
003	Fat, Crude (%)	0563	2.474	0.2272	2.469	0.2109	0.0928	143	0.03	003.14	0
003	Fat, Crude (%)	0590	2.475	0.1500	2.469	0.2109	0.0928	143	0.03	003.09	0
003	Fat, Crude (%)	0208	2.480	0.1200	2.469	0.2109	0.0928	143	0.05	003.13	0
003	Fat, Crude (%)	0682	2.480	0.0000	2.469	0.2109	0.0928	143	0.05	003.14	0
003	Fat, Crude (%)	0881	2.482	0.1090	2.469	0.2109	0.0928	143	0.06	003.10	0
003	Fat, Crude (%)	0202	2.485	0.0500	2.469	0.2109	0.0928	143	0.08	003.14	0
003	Fat, Crude (%)	0169	2.490	0.0000	2.469	0.2109	0.0928	143	0.10	003.06	0
003	Fat, Crude (%)	0870	2.494	0.1295	2.469	0.2109	0.0928	143	0.12	003.09	0
003	Fat, Crude (%)	0226	2.495	0.0900	2.469	0.2109	0.0928	143	0.12	003.09	0
003	Fat, Crude (%)	0689	2.500	0.0000	2.469	0.2109	0.0928	143	0.15	003.06	0
003	Fat, Crude (%)	0884	2.500	0.2000	2.469	0.2109	0.0928	143	0.15	003.06	0
003	Fat, Crude (%)	2319	2.500	0.0000	2.469	0.2109	0.0928	143	0.15	003.10	0
003	Fat, Crude (%)	0298	2.500	0.0200	2.469	0.2109	0.0928	143	0.15	003.14	0
003	Fat, Crude (%)	2295	2.515	0.0100	2.469	0.2109	0.0928	143	0.22	003.06	0
003	Fat, Crude (%)	0229	2.515	0.1300	2.469	0.2109	0.0928	143	0.22	003.06	0
003	Fat, Crude (%)	0893	2.520	0.0400	2.469	0.2109	0.0928	143	0.24	003.12	0
003	Fat, Crude (%)	0148	2.530	0.0400	2.469	0.2109	0.0928	143	0.29	003.10	0
003	Fat, Crude (%)	0968	2.530	0.0400	2.469	0.2109	0.0928	143	0.29	003.14	0
003	Fat, Crude (%)	0683	2.530	0.0600	2.469	0.2109	0.0928	143	0.29	003.00	0
003	Fat, Crude (%)	0294	2.535	0.0300	2.469	0.2109	0.0928	143	0.31	003.06	0
003	Fat, Crude (%)	0027	2.539	0.0420	2.469	0.2109	0.0928	143	0.33	003.09	0
003	Fat, Crude (%)	0083	2.540	0.0200	2.469	0.2109	0.0928	143	0.34	003.14	0
003	Fat, Crude (%)	2329	2.550	0.0200	2.469	0.2109	0.0928	143	0.38	003.00	0
003	Fat, Crude (%)	0590	2.550	0.0400	2.469	0.2109	0.0928	143	0.38	003.11	0
003	Fat, Crude (%)	0656	2.550	0.0000	2.469	0.2109	0.0928	143	0.38	003.99	0
003	Fat, Crude (%)	0100	2.565	0.0500	2.469	0.2109	0.0928	143	0.46	003.10	0
003	Fat, Crude (%)	2109	2.570	0.0600	2.469	0.2109	0.0928	143	0.48	003.00	0
003	Fat, Crude (%)	0353	2.575	0.0100	2.469	0.2109	0.0928	143	0.50	003.09	0
003	Fat, Crude (%)	0682	2.580	0.0000	2.469	0.2109	0.0928	143	0.53	003.06	0
003	Fat, Crude (%)	0009	2.585	0.1500	2.469	0.2109	0.0928	143	0.55	003.14	0
003	Fat, Crude (%)	0685	2.595	0.0100	2.469	0.2109	0.0928	143	0.60	003.09	0
003	Fat, Crude (%)	0511	2.595	0.1300	2.469	0.2109	0.0928	143	0.60	003.14	0
003	Fat, Crude (%)	0644	2.599	0.0860	2.469	0.2109	0.0928	143	0.62	003.06	0
003	Fat, Crude (%)	0574	2.610	0.1200	2.469	0.2109	0.0928	143	0.67	003.14	0
003	Fat, Crude (%)	2192	2.610	0.2600	2.469	0.2109	0.0928	143	0.67	003.00	0
003	Fat, Crude (%)	0553	2.610	0.0200	2.469	0.2109	0.0928	143	0.67	003.11	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
003	Fat, Crude (%)	0175	2.625	0.0100	2.469	0.2109	0.0928	143	0.74	003.00	0
003	Fat, Crude (%)	0190	2.630	0.0200	2.469	0.2109	0.0928	143	0.76	003.14	0
003	Fat, Crude (%)	2144	2.640	0.1000	2.469	0.2109	0.0928	143	0.81	003.00	0
003	Fat, Crude (%)	0074	2.640	0.3600	2.469	0.2109	0.0928	143	0.81	003.14	0
003	Fat, Crude (%)	0619	2.660	0.0800	2.469	0.2109	0.0928	143	0.91	003.14	0
003	Fat, Crude (%)	0345	2.665	0.0700	2.469	0.2109	0.0928	143	0.93	003.00	0
003	Fat, Crude (%)	0004	2.665	0.2900	2.469	0.2109	0.0928	143	0.93	003.09	0
003	Fat, Crude (%)	0297	2.685	0.4300	2.469	0.2109	0.0928	143	1.03	003.06	0
003	Fat, Crude (%)	0171	2.685	0.1700	2.469	0.2109	0.0928	143	1.03	003.12	0
003	Fat, Crude (%)	2188	2.690	0.0400	2.469	0.2109	0.0928	143	1.05	003.10	0
003	Fat, Crude (%)	0885	2.712	0.0373	2.469	0.2109	0.0928	143	1.16	003.00	0
003	Fat, Crude (%)	0066	2.721	0.2400	2.469	0.2109	0.0928	143	1.20	003.14	0
003	Fat, Crude (%)	0004	2.735	0.1500	2.469	0.2109	0.0928	143	1.26	003.14	0
003	Fat, Crude (%)	2302	2.740	0.0000	2.469	0.2109	0.0928	143	1.29	003.11	0
003	Fat, Crude (%)	0297	2.750	0.0000	2.469	0.2109	0.0928	143	1.33	003.11	0
003	Fat, Crude (%)	0278	2.750	0.1000	2.469	0.2109	0.0928	143	1.33	003.14	0
003	Fat, Crude (%)	0618	2.777	0.0270	2.469	0.2109	0.0928	143	1.46	003.10	0
003	Fat, Crude (%)	0889	2.785	0.3500	2.469	0.2109	0.0928	143	1.50	003.10	0
003	Fat, Crude (%)	0011	2.800	0.0000	2.469	0.2109	0.0928	143	1.57	003.11	0
003	Fat, Crude (%)	0089	2.805	0.0300	2.469	0.2109	0.0928	143	1.59	003.14	0
003	Fat, Crude (%)	0309	2.812	0.1002	2.469	0.2109	0.0928	143	1.63	003.00	0
003	Fat, Crude (%)	2076	2.827	0.0040	2.469	0.2109	0.0928	143	1.70	003.14	0
003	Fat, Crude (%)	0687	2.850	0.1000	2.469	0.2109	0.0928	143	1.81	003.06	0
003	Fat, Crude (%)	0164	2.865	0.0900	2.469	0.2109	0.0928	143	1.88	003.01	0
003	Fat, Crude (%)	0870	2.905	0.1176	2.469	0.2109	0.0928	143	2.07	003.13	0
003	Fat, Crude (%)	0148	2.950	0.1000	2.469	0.2109	0.0928	143	2.28	003.11	0
003	Fat, Crude (%)	2199	3.115	0.1100	2.469	0.2109	0.0928	143	3.06	003.12	0
003	Fat, Crude (%)	0852	3.165	0.0100	2.469	0.2109	0.0928	143	3.30	003.11	0
003	Fat, Crude (%)	0843	3.725	0.5300	2.469	0.2109	0.0928	143	5.96	003.14	0
003	Fat, Crude (%)	0520	4.060	0.0400	2.469	0.2109	0.0928	143	7.55	003.14	0
003	Fat, Crude (%)	0853	4.200	0.2000	2.469	0.2109	0.0928	143	8.21	003.14	0
003	Fat, Crude (%)	0003	2.490	0.5800	2.469	0.2109	0.0928	143	0.10	003.14	1
003	Fat, Crude (%)	2200	2.675	0.6500	2.469	0.2109	0.0928	143	0.98	003.12	1
003	Fat, Crude (%)	2062	2.806	0.9989	2.469	0.2109	0.0928	143	1.60	003.12	1
003	Fat, Crude (%)	0049	2.845	0.6500	2.469	0.2109	0.0928	143	1.78	003.14	1
004	Fiber, Crude (%)	0889	1.953	0.0450	3.344	0.3854	0.1754	118	-3.61	004.11	0
004	Fiber, Crude (%)	0294	2.050	0.1000	3.344	0.3854	0.1754	118	-3.36	004.07	0
004	Fiber, Crude (%)	0265	2.325	0.1700	3.344	0.3854	0.1754	118	-2.64	004.07	0
004	Fiber, Crude (%)	2089	2.415	0.3100	3.344	0.3854	0.1754	118	-2.41	004.03	0
004	Fiber, Crude (%)	0202	2.570	0.2200	3.344	0.3854	0.1754	118	-2.01	004.07	0
004	Fiber, Crude (%)	0366	2.600	1.000	3.344	0.3854	0.1754	118	-1.93	004.01	0
004	Fiber, Crude (%)	0682	2.600	0.0000	3.344	0.3854	0.1754	118	-1.93	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
004	Fiber, Crude (%)	0042	2.755	0.9700	3.344	0.3854	0.1754	118	-1.53	004.07	0
004	Fiber, Crude (%)	0921	2.780	0.3400	3.344	0.3854	0.1754	118	-1.46	004.03	0
004	Fiber, Crude (%)	2319	2.800	0.0000	3.344	0.3854	0.1754	118	-1.41	004.06	0
004	Fiber, Crude (%)	0619	2.800	0.1400	3.344	0.3854	0.1754	118	-1.41	004.07	0
004	Fiber, Crude (%)	0968	2.855	0.1700	3.344	0.3854	0.1754	118	-1.27	004.07	0
004	Fiber, Crude (%)	0708	2.870	0.0800	3.344	0.3854	0.1754	118	-1.23	004.07	0
004	Fiber, Crude (%)	2193	2.930	0.0200	3.344	0.3854	0.1754	118	-1.07	004.07	0
004	Fiber, Crude (%)	0589	2.940	0.8000	3.344	0.3854	0.1754	118	-1.05	004.07	0
004	Fiber, Crude (%)	0190	2.950	0.1000	3.344	0.3854	0.1754	118	-1.02	004.07	0
004	Fiber, Crude (%)	0726	2.960	0.0600	3.344	0.3854	0.1754	118	-1.00	004.07	0
004	Fiber, Crude (%)	0123	2.965	0.0100	3.344	0.3854	0.1754	118	-0.98	004.07	0
004	Fiber, Crude (%)	0505	2.975	0.2100	3.344	0.3854	0.1754	118	-0.96	004.07	0
004	Fiber, Crude (%)	0510	3.000	0.0000	3.344	0.3854	0.1754	118	-0.89	004.07	0
004	Fiber, Crude (%)	0004	3.025	0.3500	3.344	0.3854	0.1754	118	-0.83	004.07	0
004	Fiber, Crude (%)	0098	3.035	0.1500	3.344	0.3854	0.1754	118	-0.80	004.07	0
004	Fiber, Crude (%)	0563	3.037	0.0716	3.344	0.3854	0.1754	118	-0.79	004.07	0
004	Fiber, Crude (%)	0278	3.050	0.1000	3.344	0.3854	0.1754	118	-0.76	004.07	0
004	Fiber, Crude (%)	2022	3.060	0.1000	3.344	0.3854	0.1754	118	-0.74	004.07	0
004	Fiber, Crude (%)	0098	3.060	0.0800	3.344	0.3854	0.1754	118	-0.74	004.06	0
004	Fiber, Crude (%)	0948	3.060	0.0400	3.344	0.3854	0.1754	118	-0.74	004.07	0
004	Fiber, Crude (%)	2188	3.070	0.0400	3.344	0.3854	0.1754	118	-0.71	004.07	0
004	Fiber, Crude (%)	0169	3.075	0.1900	3.344	0.3854	0.1754	118	-0.70	004.00	0
004	Fiber, Crude (%)	0175	3.080	0.0800	3.344	0.3854	0.1754	118	-0.68	004.00	0
004	Fiber, Crude (%)	0948	3.080	0.0200	3.344	0.3854	0.1754	118	-0.68	004.06	0
004	Fiber, Crude (%)	2146	3.100	0.0800	3.344	0.3854	0.1754	118	-0.63	004.00	0
004	Fiber, Crude (%)	0692	3.100	0.0000	3.344	0.3854	0.1754	118	-0.63	004.07	0
004	Fiber, Crude (%)	0353	3.105	0.0500	3.344	0.3854	0.1754	118	-0.62	004.03	0
004	Fiber, Crude (%)	0100	3.110	0.0800	3.344	0.3854	0.1754	118	-0.61	004.07	0
004	Fiber, Crude (%)	0553	3.115	0.0900	3.344	0.3854	0.1754	118	-0.59	004.11	0
004	Fiber, Crude (%)	0643	3.150	0.0800	3.344	0.3854	0.1754	118	-0.50	004.07	0
004	Fiber, Crude (%)	0914	3.150	0.3000	3.344	0.3854	0.1754	118	-0.50	004.07	0
004	Fiber, Crude (%)	2290	3.150	0.3000	3.344	0.3854	0.1754	118	-0.50	004.11	0
004	Fiber, Crude (%)	2144	3.160	0.0400	3.344	0.3854	0.1754	118	-0.48	004.01	0
004	Fiber, Crude (%)	0009	3.160	0.0400	3.344	0.3854	0.1754	118	-0.48	004.07	0
004	Fiber, Crude (%)	0263	3.165	0.0278	3.344	0.3854	0.1754	118	-0.46	004.07	0
004	Fiber, Crude (%)	0083	3.180	0.0800	3.344	0.3854	0.1754	118	-0.42	004.07	0
004	Fiber, Crude (%)	0656	3.200	0.0000	3.344	0.3854	0.1754	118	-0.37	004.07	0
004	Fiber, Crude (%)	0683	3.200	0.1600	3.344	0.3854	0.1754	118	-0.37	004.07	0
004	Fiber, Crude (%)	0910	3.200	0.2000	3.344	0.3854	0.1754	118	-0.37	004.07	0
004	Fiber, Crude (%)	0144	3.215	0.3100	3.344	0.3854	0.1754	118	-0.33	004.07	0
004	Fiber, Crude (%)	0918	3.240	0.0200	3.344	0.3854	0.1754	118	-0.27	004.06	0
004	Fiber, Crude (%)	2146	3.240	0.2200	3.344	0.3854	0.1754	118	-0.27	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
004	Fiber, Crude (%)	0242	3.245	0.2100	3.344	0.3854	0.1754	118	-0.26	004.07	0
004	Fiber, Crude (%)	0893	3.280	0.0400	3.344	0.3854	0.1754	118	-0.16	004.06	0
004	Fiber, Crude (%)	0590	3.285	0.3300	3.344	0.3854	0.1754	118	-0.15	004.07	0
004	Fiber, Crude (%)	0689	3.300	0.0000	3.344	0.3854	0.1754	118	-0.11	004.06	0
004	Fiber, Crude (%)	0868	3.300	0.0800	3.344	0.3854	0.1754	118	-0.11	004.06	0
004	Fiber, Crude (%)	0028	3.300	0.4000	3.344	0.3854	0.1754	118	-0.11	004.07	0
004	Fiber, Crude (%)	2196	3.300	0.0000	3.344	0.3854	0.1754	118	-0.11	004.07	0
004	Fiber, Crude (%)	0035	3.315	0.1100	3.344	0.3854	0.1754	118	-0.07	004.07	0
004	Fiber, Crude (%)	0208	3.325	0.1700	3.344	0.3854	0.1754	118	-0.05	004.00	0
004	Fiber, Crude (%)	0529	3.330	0.2000	3.344	0.3854	0.1754	118	-0.04	004.07	0
004	Fiber, Crude (%)	0870	3.339	0.1023	3.344	0.3854	0.1754	118	-0.01	004.07	0
004	Fiber, Crude (%)	2302	3.345	0.5300	3.344	0.3854	0.1754	118	0.00	004.07	0
004	Fiber, Crude (%)	0876	3.350	0.3000	3.344	0.3854	0.1754	118	0.02	004.07	0
004	Fiber, Crude (%)	0536	3.390	0.0400	3.344	0.3854	0.1754	118	0.12	004.07	0
004	Fiber, Crude (%)	0298	3.400	0.2000	3.344	0.3854	0.1754	118	0.15	004.07	0
004	Fiber, Crude (%)	0581	3.400	0.2000	3.344	0.3854	0.1754	118	0.15	004.07	0
004	Fiber, Crude (%)	0941	3.415	0.0900	3.344	0.3854	0.1754	118	0.19	004.06	0
004	Fiber, Crude (%)	2181	3.440	0.2600	3.344	0.3854	0.1754	118	0.25	004.07	0
004	Fiber, Crude (%)	0164	3.450	0.1000	3.344	0.3854	0.1754	118	0.28	004.00	0
004	Fiber, Crude (%)	2246	3.450	0.0200	3.344	0.3854	0.1754	118	0.28	004.03	0
004	Fiber, Crude (%)	0148	3.450	0.1000	3.344	0.3854	0.1754	118	0.28	004.11	0
004	Fiber, Crude (%)	2058	3.450	0.0200	3.344	0.3854	0.1754	118	0.28	004.99	0
004	Fiber, Crude (%)	2076	3.462	0.3130	3.344	0.3854	0.1754	118	0.31	004.00	0
004	Fiber, Crude (%)	0881	3.465	0.0860	3.344	0.3854	0.1754	118	0.32	004.06	0
004	Fiber, Crude (%)	0027	3.468	0.0560	3.344	0.3854	0.1754	118	0.32	004.06	0
004	Fiber, Crude (%)	2062	3.474	0.1044	3.344	0.3854	0.1754	118	0.34	004.06	0
004	Fiber, Crude (%)	0878	3.475	0.0300	3.344	0.3854	0.1754	118	0.34	004.06	0
004	Fiber, Crude (%)	0309	3.476	0.0332	3.344	0.3854	0.1754	118	0.34	004.00	0
004	Fiber, Crude (%)	0171	3.485	0.2500	3.344	0.3854	0.1754	118	0.37	004.00	0
004	Fiber, Crude (%)	2109	3.495	0.0500	3.344	0.3854	0.1754	118	0.39	004.07	0
004	Fiber, Crude (%)	0610	3.500	0.1200	3.344	0.3854	0.1754	118	0.41	004.06	0
004	Fiber, Crude (%)	0520	3.500	0.0000	3.344	0.3854	0.1754	118	0.41	004.07	0
004	Fiber, Crude (%)	0183	3.513	0.1130	3.344	0.3854	0.1754	118	0.44	004.00	0
004	Fiber, Crude (%)	0881	3.520	0.2790	3.344	0.3854	0.1754	118	0.46	004.07	0
004	Fiber, Crude (%)	0019	3.525	0.0500	3.344	0.3854	0.1754	118	0.47	004.07	0
004	Fiber, Crude (%)	0066	3.535	0.3100	3.344	0.3854	0.1754	118	0.50	004.07	0
004	Fiber, Crude (%)	0553	3.545	0.0900	3.344	0.3854	0.1754	118	0.52	004.07	0
004	Fiber, Crude (%)	0226	3.550	0.2200	3.344	0.3854	0.1754	118	0.54	004.00	0
004	Fiber, Crude (%)	0511	3.550	0.1000	3.344	0.3854	0.1754	118	0.54	004.07	0
004	Fiber, Crude (%)	2009	3.566	0.0229	3.344	0.3854	0.1754	118	0.58	004.07	0
004	Fiber, Crude (%)	2006	3.600	0.6000	3.344	0.3854	0.1754	118	0.67	004.06	0
004	Fiber, Crude (%)	0011	3.600	0.2000	3.344	0.3854	0.1754	118	0.67	004.11	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
004	Fiber, Crude (%)	2326	3.605	0.0300	3.344	0.3854	0.1754	118	0.68	004.00	0
004	Fiber, Crude (%)	2200	3.625	0.3500	3.344	0.3854	0.1754	118	0.73	004.06	0
004	Fiber, Crude (%)	2295	3.655	0.0100	3.344	0.3854	0.1754	118	0.81	004.06	0
004	Fiber, Crude (%)	0407	3.656	0.2986	3.344	0.3854	0.1754	118	0.81	004.07	0
004	Fiber, Crude (%)	0968	3.680	0.0200	3.344	0.3854	0.1754	118	0.87	004.06	0
004	Fiber, Crude (%)	0425	3.700	0.0000	3.344	0.3854	0.1754	118	0.92	004.00	0
004	Fiber, Crude (%)	0939	3.705	0.2300	3.344	0.3854	0.1754	118	0.94	004.07	0
004	Fiber, Crude (%)	2073	3.715	0.0300	3.344	0.3854	0.1754	118	0.96	004.00	0
004	Fiber, Crude (%)	0897	3.720	0.0200	3.344	0.3854	0.1754	118	0.98	004.06	0
004	Fiber, Crude (%)	0638	3.745	0.0000	3.344	0.3854	0.1754	118	1.04	004.06	0
004	Fiber, Crude (%)	0644	3.754	0.6550	3.344	0.3854	0.1754	118	1.06	004.07	0
004	Fiber, Crude (%)	0074	3.780	0.2400	3.344	0.3854	0.1754	118	1.13	004.07	0
004	Fiber, Crude (%)	2302	3.830	0.0000	3.344	0.3854	0.1754	118	1.26	004.11	0
004	Fiber, Crude (%)	0089	3.835	0.0700	3.344	0.3854	0.1754	118	1.28	004.07	0
004	Fiber, Crude (%)	0034	3.845	0.3900	3.344	0.3854	0.1754	118	1.30	004.07	0
004	Fiber, Crude (%)	0885	3.869	0.1251	3.344	0.3854	0.1754	118	1.36	004.06	0
004	Fiber, Crude (%)	0345	3.950	0.3000	3.344	0.3854	0.1754	118	1.57	004.00	0
004	Fiber, Crude (%)	0884	3.950	0.1000	3.344	0.3854	0.1754	118	1.57	004.00	0
004	Fiber, Crude (%)	2192	4.010	0.3200	3.344	0.3854	0.1754	118	1.73	004.03	0
004	Fiber, Crude (%)	0728	4.025	0.1900	3.344	0.3854	0.1754	118	1.77	004.06	0
004	Fiber, Crude (%)	0598	4.080	0.2200	3.344	0.3854	0.1754	118	1.91	004.07	0
004	Fiber, Crude (%)	2291	4.095	0.1300	3.344	0.3854	0.1754	118	1.95	004.06	0
004	Fiber, Crude (%)	0693	4.149	0.8940	3.344	0.3854	0.1754	118	2.09	004.07	0
004	Fiber, Crude (%)	0852	4.385	0.0700	3.344	0.3854	0.1754	118	2.70	004.11	0
004	Fiber, Crude (%)	2129	4.527	0.5440	3.344	0.3854	0.1754	118	3.07	004.99	0
004	Fiber, Crude (%)	0045	5.000	0.6600	3.344	0.3854	0.1754	118	4.30	004.07	0
004	Fiber, Crude (%)	0626	5.865	0.2300	3.344	0.3854	0.1754	118	6.54	004.07	0
004	Fiber, Crude (%)	0186	176.8	346.4	3.344	0.3854	0.1754	118	450.09	004.00	2
005	Ash (%)	2081	5.565	0.1900	6.184	0.2748	0.0615	111	-2.25	005.99	0
005	Ash (%)	0169	5.610	0.0200	6.184	0.2748	0.0615	111	-2.09	005.00	0
005	Ash (%)	0144	5.655	0.0300	6.184	0.2748	0.0615	111	-1.93	005.00	0
005	Ash (%)	0662	5.694	0.2249	6.184	0.2748	0.0615	111	-1.78	005.99	0
005	Ash (%)	0953	5.705	0.0300	6.184	0.2748	0.0615	111	-1.74	005.00	0
005	Ash (%)	2184	5.725	0.2500	6.184	0.2748	0.0615	111	-1.67	005.00	0
005	Ash (%)	0019	5.750	0.1000	6.184	0.2748	0.0615	111	-1.58	005.00	0
005	Ash (%)	0912	5.750	0.1000	6.184	0.2748	0.0615	111	-1.58	005.03	0
005	Ash (%)	2246	5.780	0.0000	6.184	0.2748	0.0615	111	-1.47	005.00	0
005	Ash (%)	0049	5.785	0.1500	6.184	0.2748	0.0615	111	-1.45	005.00	0
005	Ash (%)	0541	5.790	0.0200	6.184	0.2748	0.0615	111	-1.43	005.00	0
005	Ash (%)	0553	5.815	0.0900	6.184	0.2748	0.0615	111	-1.34	005.00	0
005	Ash (%)	0208	5.860	0.0400	6.184	0.2748	0.0615	111	-1.18	005.00	0
005	Ash (%)	0682	5.860	0.0000	6.184	0.2748	0.0615	111	-1.18	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
005	Ash (%)	0539	5.865	0.1900	6.184	0.2748	0.0615	111	-1.16	005.00	0
005	Ash (%)	0278	5.875	0.0100	6.184	0.2748	0.0615	111	-1.13	005.00	0
005	Ash (%)	2022	5.875	0.0100	6.184	0.2748	0.0615	111	-1.13	005.00	0
005	Ash (%)	0357	5.900	0.0000	6.184	0.2748	0.0615	111	-1.03	005.00	0
005	Ash (%)	0650	5.905	0.0300	6.184	0.2748	0.0615	111	-1.02	005.00	0
005	Ash (%)	2150	5.910	0.2000	6.184	0.2748	0.0615	111	-1.00	005.00	0
005	Ash (%)	0726	5.910	0.0000	6.184	0.2748	0.0615	111	-1.00	005.05	0
005	Ash (%)	0229	5.925	0.1500	6.184	0.2748	0.0615	111	-0.94	005.00	0
005	Ash (%)	0881	5.925	0.0300	6.184	0.2748	0.0615	111	-0.94	005.00	0
005	Ash (%)	2290	5.950	0.3000	6.184	0.2748	0.0615	111	-0.85	005.11	0
005	Ash (%)	0853	5.950	0.1000	6.184	0.2748	0.0615	111	-0.85	005.00	0
005	Ash (%)	0918	5.950	0.0000	6.184	0.2748	0.0615	111	-0.85	005.00	0
005	Ash (%)	2200	5.950	0.1000	6.184	0.2748	0.0615	111	-0.85	005.00	0
005	Ash (%)	0870	5.966	0.1087	6.184	0.2748	0.0615	111	-0.80	005.00	0
005	Ash (%)	0175	5.970	0.0400	6.184	0.2748	0.0615	111	-0.78	005.00	0
005	Ash (%)	0574	5.995	0.2100	6.184	0.2748	0.0615	111	-0.69	005.00	0
005	Ash (%)	0589	5.995	0.0900	6.184	0.2748	0.0615	111	-0.69	005.00	0
005	Ash (%)	0598	5.995	0.0300	6.184	0.2748	0.0615	111	-0.69	005.00	0
005	Ash (%)	2073	6.015	0.0500	6.184	0.2748	0.0615	111	-0.62	005.00	0
005	Ash (%)	0520	6.025	0.1300	6.184	0.2748	0.0615	111	-0.58	005.00	0
005	Ash (%)	2302	6.040	0.3400	6.184	0.2748	0.0615	111	-0.53	005.00	0
005	Ash (%)	0297	6.050	0.1000	6.184	0.2748	0.0615	111	-0.49	005.00	0
005	Ash (%)	0728	6.057	0.0630	6.184	0.2748	0.0615	111	-0.47	005.99	0
005	Ash (%)	0358	6.060	0.1000	6.184	0.2748	0.0615	111	-0.45	005.00	0
005	Ash (%)	0108	6.075	0.1300	6.184	0.2748	0.0615	111	-0.40	005.00	0
005	Ash (%)	0142	6.085	0.0500	6.184	0.2748	0.0615	111	-0.36	005.00	0
005	Ash (%)	0294	6.100	0.0000	6.184	0.2748	0.0615	111	-0.31	005.05	0
005	Ash (%)	0693	6.116	0.1210	6.184	0.2748	0.0615	111	-0.25	005.00	0
005	Ash (%)	0186	6.120	0.0600	6.184	0.2748	0.0615	111	-0.23	005.00	0
005	Ash (%)	0263	6.121	0.0054	6.184	0.2748	0.0615	111	-0.23	005.05	0
005	Ash (%)	0123	6.155	0.0700	6.184	0.2748	0.0615	111	-0.11	005.00	0
005	Ash (%)	0353	6.155	0.0100	6.184	0.2748	0.0615	111	-0.11	005.00	0
005	Ash (%)	0045	6.165	0.0100	6.184	0.2748	0.0615	111	-0.07	005.00	0
005	Ash (%)	2295	6.165	0.0100	6.184	0.2748	0.0615	111	-0.07	005.05	0
005	Ash (%)	0893	6.170	0.0600	6.184	0.2748	0.0615	111	-0.05	005.05	0
005	Ash (%)	0042	6.185	0.0300	6.184	0.2748	0.0615	111	0.00	005.00	0
005	Ash (%)	2193	6.195	0.0100	6.184	0.2748	0.0615	111	0.04	005.00	0
005	Ash (%)	0897	6.195	0.0300	6.184	0.2748	0.0615	111	0.04	005.05	0
005	Ash (%)	2199	6.200	0.2000	6.184	0.2748	0.0615	111	0.06	005.00	0
005	Ash (%)	0968	6.200	0.0200	6.184	0.2748	0.0615	111	0.06	005.00	0
005	Ash (%)	0098	6.230	0.0400	6.184	0.2748	0.0615	111	0.17	005.00	0
005	Ash (%)	0148	6.230	0.0400	6.184	0.2748	0.0615	111	0.17	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
005	Ash (%)	0683	6.230	0.1000	6.184	0.2748	0.0615	111	0.17	005.05	0
005	Ash (%)	0536	6.230	0.0600	6.184	0.2748	0.0615	111	0.17	005.99	0
005	Ash (%)	0889	6.245	0.1100	6.184	0.2748	0.0615	111	0.22	005.05	0
005	Ash (%)	0407	6.246	0.0454	6.184	0.2748	0.0615	111	0.22	005.00	0
005	Ash (%)	0265	6.250	0.1000	6.184	0.2748	0.0615	111	0.24	005.00	0
005	Ash (%)	2196	6.250	0.0000	6.184	0.2748	0.0615	111	0.24	005.00	0
005	Ash (%)	0298	6.260	0.0200	6.184	0.2748	0.0615	111	0.28	005.00	0
005	Ash (%)	0083	6.265	0.0300	6.184	0.2748	0.0615	111	0.29	005.00	0
005	Ash (%)	2192	6.270	0.0600	6.184	0.2748	0.0615	111	0.31	005.05	0
005	Ash (%)	2076	6.292	0.0532	6.184	0.2748	0.0615	111	0.39	005.00	0
005	Ash (%)	0692	6.300	0.0000	6.184	0.2748	0.0615	111	0.42	005.05	0
005	Ash (%)	0884	6.300	0.0000	6.184	0.2748	0.0615	111	0.42	005.99	0
005	Ash (%)	0366	6.300	0.2000	6.184	0.2748	0.0615	111	0.42	005.00	0
005	Ash (%)	2006	6.305	0.0100	6.184	0.2748	0.0615	111	0.44	005.00	0
005	Ash (%)	2181	6.305	0.0300	6.184	0.2748	0.0615	111	0.44	005.00	0
005	Ash (%)	0868	6.305	0.0100	6.184	0.2748	0.0615	111	0.44	005.05	0
005	Ash (%)	2109	6.310	0.0800	6.184	0.2748	0.0615	111	0.46	005.05	0
005	Ash (%)	0183	6.318	0.0560	6.184	0.2748	0.0615	111	0.49	005.00	0
005	Ash (%)	0689	6.320	0.0200	6.184	0.2748	0.0615	111	0.49	005.00	0
005	Ash (%)	0242	6.335	0.0100	6.184	0.2748	0.0615	111	0.55	005.00	0
005	Ash (%)	0309	6.347	0.0460	6.184	0.2748	0.0615	111	0.59	005.00	0
005	Ash (%)	2009	6.359	0.0114	6.184	0.2748	0.0615	111	0.64	005.05	0
005	Ash (%)	2291	6.360	0.0800	6.184	0.2748	0.0615	111	0.64	005.05	0
005	Ash (%)	0878	6.370	0.0200	6.184	0.2748	0.0615	111	0.68	005.05	0
005	Ash (%)	0164	6.375	0.0300	6.184	0.2748	0.0615	111	0.69	005.00	0
005	Ash (%)	0425	6.375	0.1500	6.184	0.2748	0.0615	111	0.69	005.00	0
005	Ash (%)	0948	6.375	0.1300	6.184	0.2748	0.0615	111	0.69	005.00	0
005	Ash (%)	2089	6.380	0.0000	6.184	0.2748	0.0615	111	0.71	005.00	0
005	Ash (%)	0202	6.380	0.0000	6.184	0.2748	0.0615	111	0.71	005.99	0
005	Ash (%)	0618	6.385	0.0351	6.184	0.2748	0.0615	111	0.73	005.00	0
005	Ash (%)	0638	6.395	0.0100	6.184	0.2748	0.0615	111	0.77	005.05	0
005	Ash (%)	0345	6.400	0.0200	6.184	0.2748	0.0615	111	0.78	005.00	0
005	Ash (%)	0510	6.400	0.0000	6.184	0.2748	0.0615	111	0.78	005.00	0
005	Ash (%)	0590	6.405	0.0100	6.184	0.2748	0.0615	111	0.80	005.05	0
005	Ash (%)	0939	6.405	0.0100	6.184	0.2748	0.0615	111	0.80	005.05	0
005	Ash (%)	0563	6.419	0.1028	6.184	0.2748	0.0615	111	0.85	005.00	0
005	Ash (%)	0581	6.420	0.0200	6.184	0.2748	0.0615	111	0.86	005.00	0
005	Ash (%)	0885	6.431	0.0224	6.184	0.2748	0.0615	111	0.90	005.05	0
005	Ash (%)	0171	6.435	0.0500	6.184	0.2748	0.0615	111	0.91	005.00	0
005	Ash (%)	2188	6.445	0.0500	6.184	0.2748	0.0615	111	0.95	005.00	0
005	Ash (%)	0910	6.450	0.1000	6.184	0.2748	0.0615	111	0.97	005.99	0
005	Ash (%)	0644	6.453	0.0490	6.184	0.2748	0.0615	111	0.98	005.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
005	Ash (%)	0921	6.470	0.0200	6.184	0.2748	0.0615	111	1.04	005.05	0
005	Ash (%)	2146	6.485	0.0100	6.184	0.2748	0.0615	111	1.09	005.05	0
005	Ash (%)	0652	6.500	0.0000	6.184	0.2748	0.0615	111	1.15	005.99	0
005	Ash (%)	0034	6.509	0.0320	6.184	0.2748	0.0615	111	1.18	005.05	0
005	Ash (%)	0226	6.530	0.1200	6.184	0.2748	0.0615	111	1.26	005.00	0
005	Ash (%)	0619	6.535	0.0500	6.184	0.2748	0.0615	111	1.28	005.05	0
005	Ash (%)	2062	6.540	0.0244	6.184	0.2748	0.0615	111	1.29	005.05	0
005	Ash (%)	0643	6.565	0.0100	6.184	0.2748	0.0615	111	1.39	005.00	0
005	Ash (%)	2246	6.565	0.0500	6.184	0.2748	0.0615	111	1.39	005.05	0
005	Ash (%)	0100	6.585	0.0300	6.184	0.2748	0.0615	111	1.46	005.00	0
005	Ash (%)	0685	6.645	0.0100	6.184	0.2748	0.0615	111	1.68	005.05	0
005	Ash (%)	0148	6.700	0.0000	6.184	0.2748	0.0615	111	1.88	005.11	0
005	Ash (%)	0852	7.905	0.0100	6.184	0.2748	0.0615	111	6.26	005.11	0
005	Ash (%)	2129	6.344	0.7110	6.184	0.2748	0.0615	111	0.58	005.05	1
005	Ash (%)	0297	7.670	0.3800	6.184	0.2748	0.0615	111	5.41	005.11	1
005	Ash (%)	0889	9.128	0.0150	6.184	0.2748	0.0615	111	10.71	005.11	2
006	Total Sugars (%)	0407	4.318	0.0020	5.186	0.8243	0.1653	6	-1.05	006.00	0
006	Total Sugars (%)	0910	4.785	0.1700	5.186	0.8243	0.1653	6	-0.49	006.00	0
006	Total Sugars (%)	2290	4.800	0.4000	5.186	0.8243	0.1653	6	-0.47	006.99	0
006	Total Sugars (%)	0941	5.110	0.2200	5.186	0.8243	0.1653	6	-0.09	006.03	0
006	Total Sugars (%)	2192	5.850	0.1000	5.186	0.8243	0.1653	6	0.81	006.03	0
006	Total Sugars (%)	0912	6.250	0.1000	5.186	0.8243	0.1653	6	1.29	006.99	0
008	Fiber, Acid Detergent (%)	2129	3.649	0.0930	4.504	0.3815	0.1914	58	-2.24	008.08	0
008	Fiber, Acid Detergent (%)	2143	3.795	0.0700	4.504	0.3815	0.1914	58	-1.86	008.08	0
008	Fiber, Acid Detergent (%)	0042	3.910	0.3200	4.504	0.3815	0.1914	58	-1.56	008.08	0
008	Fiber, Acid Detergent (%)	0004	3.925	0.1500	4.504	0.3815	0.1914	58	-1.52	008.08	0
008	Fiber, Acid Detergent (%)	2302	3.925	0.1300	4.504	0.3815	0.1914	58	-1.52	008.08	0
008	Fiber, Acid Detergent (%)	0968	3.990	0.1600	4.504	0.3815	0.1914	58	-1.35	008.08	0
008	Fiber, Acid Detergent (%)	0590	4.005	0.2100	4.504	0.3815	0.1914	58	-1.31	008.08	0
008	Fiber, Acid Detergent (%)	0948	4.030	0.1200	4.504	0.3815	0.1914	58	-1.24	008.08	0
008	Fiber, Acid Detergent (%)	0939	4.045	0.1300	4.504	0.3815	0.1914	58	-1.20	008.08	0
008	Fiber, Acid Detergent (%)	0619	4.070	0.4000	4.504	0.3815	0.1914	58	-1.14	008.08	0
008	Fiber, Acid Detergent (%)	0948	4.085	0.0500	4.504	0.3815	0.1914	58	-1.10	008.02	0
008	Fiber, Acid Detergent (%)	0358	4.145	0.3900	4.504	0.3815	0.1914	58	-0.94	008.08	0
008	Fiber, Acid Detergent (%)	0098	4.230	0.0600	4.504	0.3815	0.1914	58	-0.72	008.02	0
008	Fiber, Acid Detergent (%)	0027	4.249	0.0190	4.504	0.3815	0.1914	58	-0.67	008.99	0
008	Fiber, Acid Detergent (%)	0294	4.250	0.1000	4.504	0.3815	0.1914	58	-0.66	008.08	0
008	Fiber, Acid Detergent (%)	0037	4.320	0.2400	4.504	0.3815	0.1914	58	-0.48	008.08	0
008	Fiber, Acid Detergent (%)	0656	4.330	0.0000	4.504	0.3815	0.1914	58	-0.46	008.08	0
008	Fiber, Acid Detergent (%)	0941	4.330	0.1400	4.504	0.3815	0.1914	58	-0.46	008.99	0
008	Fiber, Acid Detergent (%)	0510	4.345	0.0500	4.504	0.3815	0.1914	58	-0.42	008.08	0
008	Fiber, Acid Detergent (%)	0297	4.380	0.4400	4.504	0.3815	0.1914	58	-0.32	008.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
008	Fiber, Acid Detergent (%)	0511	4.400	0.2000	4.504	0.3815	0.1914	58	-0.27	008.99	0
008	Fiber, Acid Detergent (%)	0912	4.405	0.3100	4.504	0.3815	0.1914	58	-0.26	008.08	0
008	Fiber, Acid Detergent (%)	2193	4.430	0.2800	4.504	0.3815	0.1914	58	-0.19	008.08	0
008	Fiber, Acid Detergent (%)	0083	4.455	0.0500	4.504	0.3815	0.1914	58	-0.13	008.08	0
008	Fiber, Acid Detergent (%)	0265	4.460	0.8400	4.504	0.3815	0.1914	58	-0.11	008.05	0
008	Fiber, Acid Detergent (%)	0309	4.463	0.0260	4.504	0.3815	0.1914	58	-0.11	008.02	0
008	Fiber, Acid Detergent (%)	0885	4.472	0.0115	4.504	0.3815	0.1914	58	-0.08	008.02	0
008	Fiber, Acid Detergent (%)	0918	4.480	0.0800	4.504	0.3815	0.1914	58	-0.06	008.08	0
008	Fiber, Acid Detergent (%)	0726	4.490	0.0800	4.504	0.3815	0.1914	58	-0.04	008.08	0
008	Fiber, Acid Detergent (%)	0581	4.550	0.1000	4.504	0.3815	0.1914	58	0.12	008.08	0
008	Fiber, Acid Detergent (%)	2146	4.590	0.6200	4.504	0.3815	0.1914	58	0.23	008.08	0
008	Fiber, Acid Detergent (%)	0357	4.600	0.0000	4.504	0.3815	0.1914	58	0.25	008.08	0
008	Fiber, Acid Detergent (%)	0345	4.625	0.1500	4.504	0.3815	0.1914	58	0.32	008.02	0
008	Fiber, Acid Detergent (%)	0263	4.633	0.0156	4.504	0.3815	0.1914	58	0.34	008.08	0
008	Fiber, Acid Detergent (%)	0889	4.640	0.0000	4.504	0.3815	0.1914	58	0.36	008.08	0
008	Fiber, Acid Detergent (%)	0190	4.650	0.2400	4.504	0.3815	0.1914	58	0.38	008.08	0
008	Fiber, Acid Detergent (%)	0278	4.650	0.1000	4.504	0.3815	0.1914	58	0.38	008.08	0
008	Fiber, Acid Detergent (%)	0226	4.655	0.0900	4.504	0.3815	0.1914	58	0.40	008.02	0
008	Fiber, Acid Detergent (%)	0878	4.675	0.1300	4.504	0.3815	0.1914	58	0.45	008.08	0
008	Fiber, Acid Detergent (%)	0693	4.682	0.5920	4.504	0.3815	0.1914	58	0.47	008.08	0
008	Fiber, Acid Detergent (%)	0148	4.720	0.0400	4.504	0.3815	0.1914	58	0.57	008.08	0
008	Fiber, Acid Detergent (%)	0045	4.735	0.1700	4.504	0.3815	0.1914	58	0.61	008.08	0
008	Fiber, Acid Detergent (%)	2192	4.745	0.7700	4.504	0.3815	0.1914	58	0.63	008.02	0
008	Fiber, Acid Detergent (%)	0910	4.745	0.1100	4.504	0.3815	0.1914	58	0.63	008.08	0
008	Fiber, Acid Detergent (%)	0728	4.750	0.5000	4.504	0.3815	0.1914	58	0.65	008.02	0
008	Fiber, Acid Detergent (%)	0066	4.753	0.2030	4.504	0.3815	0.1914	58	0.65	008.08	0
008	Fiber, Acid Detergent (%)	0610	4.785	0.1500	4.504	0.3815	0.1914	58	0.74	008.99	0
008	Fiber, Acid Detergent (%)	0870	4.789	0.1603	4.504	0.3815	0.1914	58	0.75	008.08	0
008	Fiber, Acid Detergent (%)	0001	4.816	0.4010	4.504	0.3815	0.1914	58	0.82	008.08	0
008	Fiber, Acid Detergent (%)	0148	4.820	0.0600	4.504	0.3815	0.1914	58	0.83	008.02	0
008	Fiber, Acid Detergent (%)	2196	4.850	0.0000	4.504	0.3815	0.1914	58	0.91	008.08	0
008	Fiber, Acid Detergent (%)	0536	4.915	0.0100	4.504	0.3815	0.1914	58	1.08	008.08	0
008	Fiber, Acid Detergent (%)	0035	5.005	0.0100	4.504	0.3815	0.1914	58	1.31	008.08	0
008	Fiber, Acid Detergent (%)	0353	5.040	0.1400	4.504	0.3815	0.1914	58	1.41	008.02	0
008	Fiber, Acid Detergent (%)	2290	5.050	0.3000	4.504	0.3815	0.1914	58	1.43	008.99	0
008	Fiber, Acid Detergent (%)	0164	5.100	0.4000	4.504	0.3815	0.1914	58	1.56	008.08	0
008	Fiber, Acid Detergent (%)	0407	5.287	0.4922	4.504	0.3815	0.1914	58	2.05	008.08	0
008	Fiber, Acid Detergent (%)	0689	5.500	0.0000	4.504	0.3815	0.1914	58	2.61	008.02	0
008	Fiber, Acid Detergent (%)	0049	3.905	0.9900	4.504	0.3815	0.1914	58	-1.57	008.08	1
009	Fiber, Neutral Detergent (%)	0910	8.800	0.2000	11.03	0.8010	0.3156	57	-2.78	009.07	0
009	Fiber, Neutral Detergent (%)	0049	9.845	0.2300	11.03	0.8010	0.3156	57	-1.48	009.09	0
009	Fiber, Neutral Detergent (%)	0598	10.04	0.0700	11.03	0.8010	0.3156	57	-1.24	009.09	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
009	Fiber, Neutral Detergent (%)	0098	10.04	0.5200	11.03	0.8010	0.3156	57	-1.23	009.07	0
009	Fiber, Neutral Detergent (%)	0358	10.05	0.5700	11.03	0.8010	0.3156	57	-1.23	009.09	0
009	Fiber, Neutral Detergent (%)	0353	10.11	0.7000	11.03	0.8010	0.3156	57	-1.14	009.07	0
009	Fiber, Neutral Detergent (%)	0619	10.25	0.1000	11.03	0.8010	0.3156	57	-0.97	009.09	0
009	Fiber, Neutral Detergent (%)	0881	10.28	0.0100	11.03	0.8010	0.3156	57	-0.94	009.09	0
009	Fiber, Neutral Detergent (%)	0939	10.29	0.6200	11.03	0.8010	0.3156	57	-0.92	009.09	0
009	Fiber, Neutral Detergent (%)	2022	10.32	0.2400	11.03	0.8010	0.3156	57	-0.88	009.09	0
009	Fiber, Neutral Detergent (%)	0912	10.34	0.2700	11.03	0.8010	0.3156	57	-0.86	009.09	0
009	Fiber, Neutral Detergent (%)	2143	10.36	0.1000	11.03	0.8010	0.3156	57	-0.83	009.09	0
009	Fiber, Neutral Detergent (%)	0536	10.39	0.0700	11.03	0.8010	0.3156	57	-0.80	009.09	0
009	Fiber, Neutral Detergent (%)	0656	10.47	0.0000	11.03	0.8010	0.3156	57	-0.70	009.09	0
009	Fiber, Neutral Detergent (%)	0693	10.48	1.022	11.03	0.8010	0.3156	57	-0.69	009.09	0
009	Fiber, Neutral Detergent (%)	0357	10.50	0.0000	11.03	0.8010	0.3156	57	-0.66	009.09	0
009	Fiber, Neutral Detergent (%)	0263	10.51	0.0840	11.03	0.8010	0.3156	57	-0.64	009.09	0
009	Fiber, Neutral Detergent (%)	0042	10.60	0.6000	11.03	0.8010	0.3156	57	-0.53	009.09	0
009	Fiber, Neutral Detergent (%)	0037	10.63	0.1500	11.03	0.8010	0.3156	57	-0.50	009.09	0
009	Fiber, Neutral Detergent (%)	0278	10.65	0.1000	11.03	0.8010	0.3156	57	-0.47	009.09	0
009	Fiber, Neutral Detergent (%)	0407	10.68	0.1581	11.03	0.8010	0.3156	57	-0.43	009.09	0
009	Fiber, Neutral Detergent (%)	0948	10.71	0.0800	11.03	0.8010	0.3156	57	-0.40	009.07	0
009	Fiber, Neutral Detergent (%)	0148	10.71	0.1400	11.03	0.8010	0.3156	57	-0.40	009.09	0
009	Fiber, Neutral Detergent (%)	0581	10.81	0.2700	11.03	0.8010	0.3156	57	-0.28	009.09	0
009	Fiber, Neutral Detergent (%)	0948	10.82	0.1600	11.03	0.8010	0.3156	57	-0.26	009.09	0
009	Fiber, Neutral Detergent (%)	0726	10.85	0.1400	11.03	0.8010	0.3156	57	-0.22	009.09	0
009	Fiber, Neutral Detergent (%)	0297	10.89	0.9400	11.03	0.8010	0.3156	57	-0.17	009.09	0
009	Fiber, Neutral Detergent (%)	0889	10.91	0.1300	11.03	0.8010	0.3156	57	-0.15	009.09	0
009	Fiber, Neutral Detergent (%)	0345	10.93	0.2600	11.03	0.8010	0.3156	57	-0.12	009.07	0
009	Fiber, Neutral Detergent (%)	2302	10.97	0.1100	11.03	0.8010	0.3156	57	-0.08	009.09	0
009	Fiber, Neutral Detergent (%)	0035	11.01	0.3400	11.03	0.8010	0.3156	57	-0.02	009.09	0
009	Fiber, Neutral Detergent (%)	2146	11.02	1.460	11.03	0.8010	0.3156	57	-0.01	009.09	0
009	Fiber, Neutral Detergent (%)	0226	11.08	0.0100	11.03	0.8010	0.3156	57	0.06	009.07	0
009	Fiber, Neutral Detergent (%)	0066	11.09	0.4090	11.03	0.8010	0.3156	57	0.08	009.09	0
009	Fiber, Neutral Detergent (%)	0265	11.10	0.8000	11.03	0.8010	0.3156	57	0.09	009.09	0
009	Fiber, Neutral Detergent (%)	0878	11.10	0.0600	11.03	0.8010	0.3156	57	0.09	009.09	0
009	Fiber, Neutral Detergent (%)	2129	11.12	0.5410	11.03	0.8010	0.3156	57	0.12	009.09	0
009	Fiber, Neutral Detergent (%)	0918	11.24	0.0200	11.03	0.8010	0.3156	57	0.27	009.09	0
009	Fiber, Neutral Detergent (%)	0885	11.28	0.4804	11.03	0.8010	0.3156	57	0.32	009.07	0
009	Fiber, Neutral Detergent (%)	0083	11.37	0.0300	11.03	0.8010	0.3156	57	0.42	009.09	0
009	Fiber, Neutral Detergent (%)	2196	11.47	0.0000	11.03	0.8010	0.3156	57	0.55	009.09	0
009	Fiber, Neutral Detergent (%)	0510	11.50	0.2000	11.03	0.8010	0.3156	57	0.59	009.09	0
009	Fiber, Neutral Detergent (%)	2193	11.53	0.1500	11.03	0.8010	0.3156	57	0.62	009.09	0
009	Fiber, Neutral Detergent (%)	0190	11.54	0.4400	11.03	0.8010	0.3156	57	0.64	009.09	0
009	Fiber, Neutral Detergent (%)	2290	11.60	0.4000	11.03	0.8010	0.3156	57	0.72	009.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
009	Fiber, Neutral Detergent (%)	0889	11.82	0.1500	11.03	0.8010	0.3156	57	0.98	009.99	0
009	Fiber, Neutral Detergent (%)	0045	12.00	0.2000	11.03	0.8010	0.3156	57	1.21	009.09	0
009	Fiber, Neutral Detergent (%)	0610	12.05	0.1000	11.03	0.8010	0.3156	57	1.28	009.07	0
009	Fiber, Neutral Detergent (%)	0968	12.06	0.5000	11.03	0.8010	0.3156	57	1.29	009.09	0
009	Fiber, Neutral Detergent (%)	0941	12.07	0.3200	11.03	0.8010	0.3156	57	1.30	009.99	0
009	Fiber, Neutral Detergent (%)	0294	12.10	0.0000	11.03	0.8010	0.3156	57	1.34	009.09	0
009	Fiber, Neutral Detergent (%)	0590	12.10	1.400	11.03	0.8010	0.3156	57	1.34	009.09	0
009	Fiber, Neutral Detergent (%)	0309	12.17	0.0087	11.03	0.8010	0.3156	57	1.42	009.07	0
009	Fiber, Neutral Detergent (%)	0164	12.45	0.7000	11.03	0.8010	0.3156	57	1.78	009.09	0
009	Fiber, Neutral Detergent (%)	0689	13.30	0.6000	11.03	0.8010	0.3156	57	2.84	009.07	0
009	Fiber, Neutral Detergent (%)	0870	13.39	0.6245	11.03	0.8010	0.3156	57	2.95	009.09	0
009	Fiber, Neutral Detergent (%)	0728	15.20	0.0000	11.03	0.8010	0.3156	57	5.21	009.99	0
009	Fiber, Neutral Detergent (%)	2192	13.03	1.880	11.03	0.8010	0.3156	57	2.50	009.07	1
010	Moisture (%)	0726	9.315	0.0100	9.996	0.5335	0.1670	23	-1.28	010.99	0
010	Moisture (%)	0889	9.398	0.0550	9.996	0.5335	0.1670	23	-1.12	010.11	0
010	Moisture (%)	2216	9.470	0.8200	9.996	0.5335	0.1670	23	-0.99	010.99	0
010	Moisture (%)	0692	9.535	0.3100	9.996	0.5335	0.1670	23	-0.86	010.99	0
010	Moisture (%)	0939	9.550	0.2000	9.996	0.5335	0.1670	23	-0.84	010.99	0
010	Moisture (%)	0948	9.595	0.0100	9.996	0.5335	0.1670	23	-0.75	010.99	0
010	Moisture (%)	0884	9.600	0.0000	9.996	0.5335	0.1670	23	-0.74	010.99	0
010	Moisture (%)	0964	9.621	0.1860	9.996	0.5335	0.1670	23	-0.70	010.99	0
010	Moisture (%)	2089	9.700	0.0600	9.996	0.5335	0.1670	23	-0.56	010.99	0
010	Moisture (%)	2129	9.836	0.7610	9.996	0.5335	0.1670	23	-0.30	010.99	0
010	Moisture (%)	2076	9.920	0.2800	9.996	0.5335	0.1670	23	-0.14	010.99	0
010	Moisture (%)	2150	9.955	0.0300	9.996	0.5335	0.1670	23	-0.08	010.99	0
010	Moisture (%)	2178	9.990	0.1200	9.996	0.5335	0.1670	23	-0.01	010.99	0
010	Moisture (%)	0921	10.03	0.1800	9.996	0.5335	0.1670	23	0.06	010.99	0
010	Moisture (%)	0148	10.20	0.0000	9.996	0.5335	0.1670	23	0.38	010.11	0
010	Moisture (%)	0914	10.25	0.2100	9.996	0.5335	0.1670	23	0.47	010.99	0
010	Moisture (%)	0652	10.25	0.1000	9.996	0.5335	0.1670	23	0.48	010.99	0
010	Moisture (%)	0164	10.26	0.0500	9.996	0.5335	0.1670	23	0.49	010.03	0
010	Moisture (%)	0190	10.37	0.0200	9.996	0.5335	0.1670	23	0.70	010.99	0
010	Moisture (%)	2290	10.69	0.1400	9.996	0.5335	0.1670	23	1.30	010.11	0
010	Moisture (%)	2329	10.84	0.2700	9.996	0.5335	0.1670	23	1.57	010.99	0
010	Moisture (%)	0852	10.93	0.0300	9.996	0.5335	0.1670	23	1.74	010.11	0
010	Moisture (%)	2302	11.72	0.0000	9.996	0.5335	0.1670	23	3.23	010.11	0
010	Moisture (%)	0843	10.24	1.150	9.996	0.5335	0.1670	23	0.45	010.03	1
010	Moisture (%)	2043	0.6700	0.0400	9.996	0.5335	0.1670	23	-17.48	010.99	2
011	Loss on Drying (%)	0685	8.795	0.0100	10.51	0.3315	0.0912	61	-5.16	011.01	0
011	Loss on Drying (%)	2200	8.850	0.1000	10.51	0.3315	0.0912	61	-5.00	011.01	0
011	Loss on Drying (%)	0536	9.405	0.0100	10.51	0.3315	0.0912	61	-3.32	011.01	0
011	Loss on Drying (%)	0859	9.448	0.0650	10.51	0.3315	0.0912	61	-3.19	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
011	Loss on Drying (%)	2291	9.775	0.1500	10.51	0.3315	0.0912	61	-2.21	011.99	0
011	Loss on Drying (%)	2188	9.860	0.1000	10.51	0.3315	0.0912	61	-1.95	011.01	0
011	Loss on Drying (%)	0520	10.10	0.2000	10.51	0.3315	0.0912	61	-1.23	011.01	0
011	Loss on Drying (%)	0953	10.14	0.0900	10.51	0.3315	0.0912	61	-1.12	011.01	0
011	Loss on Drying (%)	0100	10.15	0.1000	10.51	0.3315	0.0912	61	-1.07	011.01	0
011	Loss on Drying (%)	0881	10.16	0.1290	10.51	0.3315	0.0912	61	-1.05	011.01	0
011	Loss on Drying (%)	2326	10.22	0.0500	10.51	0.3315	0.0912	61	-0.88	011.01	0
011	Loss on Drying (%)	0598	10.23	0.2300	10.51	0.3315	0.0912	61	-0.85	011.01	0
011	Loss on Drying (%)	0226	10.24	0.0800	10.51	0.3315	0.0912	61	-0.80	011.01	0
011	Loss on Drying (%)	0852	10.30	0.0300	10.51	0.3315	0.0912	61	-0.64	011.99	0
011	Loss on Drying (%)	0265	10.30	0.0000	10.51	0.3315	0.0912	61	-0.62	011.01	0
011	Loss on Drying (%)	0510	10.30	0.0000	10.51	0.3315	0.0912	61	-0.62	011.01	0
011	Loss on Drying (%)	0968	10.31	0.0100	10.51	0.3315	0.0912	61	-0.61	011.01	0
011	Loss on Drying (%)	0511	10.31	0.2600	10.51	0.3315	0.0912	61	-0.59	011.01	0
011	Loss on Drying (%)	0539	10.33	0.0500	10.51	0.3315	0.0912	61	-0.55	011.01	0
011	Loss on Drying (%)	0298	10.34	0.0200	10.51	0.3315	0.0912	61	-0.50	011.01	0
011	Loss on Drying (%)	0425	10.34	0.0400	10.51	0.3315	0.0912	61	-0.50	011.01	0
011	Loss on Drying (%)	2181	10.36	0.0400	10.51	0.3315	0.0912	61	-0.44	011.01	0
011	Loss on Drying (%)	2303	10.36	0.3470	10.51	0.3315	0.0912	61	-0.43	011.99	0
011	Loss on Drying (%)	0108	10.43	0.2600	10.51	0.3315	0.0912	61	-0.23	011.01	0
011	Loss on Drying (%)	2006	10.43	0.0000	10.51	0.3315	0.0912	61	-0.23	011.01	0
011	Loss on Drying (%)	2246	10.44	0.0200	10.51	0.3315	0.0912	61	-0.20	011.01	0
011	Loss on Drying (%)	2174	10.44	0.1000	10.51	0.3315	0.0912	61	-0.20	011.01	0
011	Loss on Drying (%)	0294	10.45	0.1000	10.51	0.3315	0.0912	61	-0.17	011.01	0
011	Loss on Drying (%)	2196	10.49	0.0000	10.51	0.3315	0.0912	61	-0.05	011.01	0
011	Loss on Drying (%)	2168	10.50	0.0100	10.51	0.3315	0.0912	61	-0.03	011.01	0
011	Loss on Drying (%)	2073	10.52	0.0300	10.51	0.3315	0.0912	61	0.03	011.01	0
011	Loss on Drying (%)	2193	10.52	0.0300	10.51	0.3315	0.0912	61	0.03	011.01	0
011	Loss on Drying (%)	0309	10.54	0.0400	10.51	0.3315	0.0912	61	0.10	011.01	0
011	Loss on Drying (%)	0643	10.55	0.1000	10.51	0.3315	0.0912	61	0.13	011.01	0
011	Loss on Drying (%)	2022	10.57	0.0700	10.51	0.3315	0.0912	61	0.18	011.01	0
011	Loss on Drying (%)	0682	10.61	0.0000	10.51	0.3315	0.0912	61	0.31	011.01	0
011	Loss on Drying (%)	0728	10.62	0.0100	10.51	0.3315	0.0912	61	0.33	011.01	0
011	Loss on Drying (%)	2302	10.62	0.1300	10.51	0.3315	0.0912	61	0.33	011.01	0
011	Loss on Drying (%)	0098	10.65	0.1200	10.51	0.3315	0.0912	61	0.43	011.01	0
011	Loss on Drying (%)	0263	10.66	0.0100	10.51	0.3315	0.0912	61	0.45	011.01	0
011	Loss on Drying (%)	0171	10.67	0.0300	10.51	0.3315	0.0912	61	0.48	011.01	0
011	Loss on Drying (%)	0574	10.67	0.0700	10.51	0.3315	0.0912	61	0.48	011.01	0
011	Loss on Drying (%)	2184	10.70	0.1100	10.51	0.3315	0.0912	61	0.57	011.01	0
011	Loss on Drying (%)	0175	10.70	0.0000	10.51	0.3315	0.0912	61	0.58	011.01	0
011	Loss on Drying (%)	0529	10.70	0.0000	10.51	0.3315	0.0912	61	0.58	011.02	0
011	Loss on Drying (%)	0066	10.74	0.0760	10.51	0.3315	0.0912	61	0.72	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
011	Loss on Drying (%)	0553	10.75	0.0100	10.51	0.3315	0.0912	61	0.72	011.01	0
011	Loss on Drying (%)	0541	10.76	0.2700	10.51	0.3315	0.0912	61	0.75	011.01	0
011	Loss on Drying (%)	0242	10.78	0.2500	10.51	0.3315	0.0912	61	0.81	011.01	0
011	Loss on Drying (%)	0123	10.81	0.0200	10.51	0.3315	0.0912	61	0.92	011.01	0
011	Loss on Drying (%)	0144	10.83	0.0300	10.51	0.3315	0.0912	61	0.96	011.01	0
011	Loss on Drying (%)	0004	10.85	0.0800	10.51	0.3315	0.0912	61	1.04	011.01	0
011	Loss on Drying (%)	0164	10.88	0.0400	10.51	0.3315	0.0912	61	1.13	011.01	0
011	Loss on Drying (%)	0229	10.89	0.4800	10.51	0.3315	0.0912	61	1.16	011.01	0
011	Loss on Drying (%)	0407	10.90	0.1070	10.51	0.3315	0.0912	61	1.18	011.01	0
011	Loss on Drying (%)	0589	10.91	0.0100	10.51	0.3315	0.0912	61	1.20	011.01	0
011	Loss on Drying (%)	0870	10.92	0.0271	10.51	0.3315	0.0912	61	1.25	011.01	0
011	Loss on Drying (%)	0618	10.93	0.0400	10.51	0.3315	0.0912	61	1.28	011.01	0
011	Loss on Drying (%)	0202	11.09	0.3000	10.51	0.3315	0.0912	61	1.76	011.01	0
011	Loss on Drying (%)	0358	11.26	0.1700	10.51	0.3315	0.0912	61	2.26	011.01	0
011	Loss on Drying (%)	0843	11.46	0.3000	10.51	0.3315	0.0912	61	2.88	011.01	0
011	Loss on Drying (%)	2199	9.100	1.000	10.51	0.3315	0.0912	61	-4.24	011.01	1
011	Loss on Drying (%)	0563	1.839	0.0769	10.51	0.3315	0.0912	61	-26.14	011.01	2
012	Starch (%)	0918	34.18	0.8780	38.94	2.410	0.6360	34	-1.98	012.01	0
012	Starch (%)	0889	35.49	0.2500	38.94	2.410	0.6360	34	-1.43	012.11	0
012	Starch (%)	0910	35.80	0.2000	38.94	2.410	0.6360	34	-1.30	012.01	0
012	Starch (%)	0407	36.08	0.5232	38.94	2.410	0.6360	34	-1.19	012.03	0
012	Starch (%)	0510	36.35	0.5000	38.94	2.410	0.6360	34	-1.08	012.04	0
012	Starch (%)	0164	36.65	0.5000	38.94	2.410	0.6360	34	-0.95	012.01	0
012	Starch (%)	0297	36.84	1.390	38.94	2.410	0.6360	34	-0.87	012.03	0
012	Starch (%)	0098	37.00	0.4000	38.94	2.410	0.6360	34	-0.81	012.04	0
012	Starch (%)	0353	37.00	0.2200	38.94	2.410	0.6360	34	-0.81	012.20	0
012	Starch (%)	0948	37.51	1.500	38.94	2.410	0.6360	34	-0.59	012.04	0
012	Starch (%)	0278	37.65	0.5000	38.94	2.410	0.6360	34	-0.54	012.04	0
012	Starch (%)	0852	37.75	0.0500	38.94	2.410	0.6360	34	-0.50	012.11	0
012	Starch (%)	0148	38.07	1.330	38.94	2.410	0.6360	34	-0.36	012.01	0
012	Starch (%)	0689	38.10	0.2000	38.94	2.410	0.6360	34	-0.35	012.00	0
012	Starch (%)	0693	38.66	1.230	38.94	2.410	0.6360	34	-0.12	012.01	0
012	Starch (%)	2192	38.74	2.490	38.94	2.410	0.6360	34	-0.09	012.00	0
012	Starch (%)	0045	38.90	0.0000	38.94	2.410	0.6360	34	-0.02	012.01	0
012	Starch (%)	0619	39.00	1.200	38.94	2.410	0.6360	34	0.02	012.00	0
012	Starch (%)	0164	39.10	1.600	38.94	2.410	0.6360	34	0.07	012.00	0
012	Starch (%)	2129	39.42	0.4930	38.94	2.410	0.6360	34	0.20	012.04	0
012	Starch (%)	0683	39.60	0.5900	38.94	2.410	0.6360	34	0.27	012.00	0
012	Starch (%)	2181	39.78	0.2200	38.94	2.410	0.6360	34	0.35	012.00	0
012	Starch (%)	0910	40.15	0.5000	38.94	2.410	0.6360	34	0.50	012.00	0
012	Starch (%)	2006	40.18	0.0900	38.94	2.410	0.6360	34	0.51	012.00	0
012	Starch (%)	0878	40.25	0.3000	38.94	2.410	0.6360	34	0.54	012.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
012	Starch (%)	2193	40.64	0.2600	38.94	2.410	0.6360	34	0.70	012.00	0
012	Starch (%)	2290	40.90	0.2000	38.94	2.410	0.6360	34	0.81	012.11	0
012	Starch (%)	0297	41.15	0.5000	38.94	2.410	0.6360	34	0.92	012.11	0
012	Starch (%)	2188	41.30	0.2500	38.94	2.410	0.6360	34	0.98	012.00	0
012	Starch (%)	0889	41.44	0.0000	38.94	2.410	0.6360	34	1.04	012.03	0
012	Starch (%)	2196	41.74	1.000	38.94	2.410	0.6360	34	1.16	012.00	0
012	Starch (%)	2089	42.38	0.3600	38.94	2.410	0.6360	34	1.43	012.01	0
012	Starch (%)	0590	44.55	0.1000	38.94	2.410	0.6360	34	2.33	012.11	0
012	Starch (%)	0265	45.00	1.800	38.94	2.410	0.6360	34	2.51	012.01	0
012	Starch (%)	0870	3.498	0.0198	38.94	2.410	0.6360	34	-14.71	012.01	2
013	Fat, Acid Pretreat (%)	0618	1.764	0.4822	3.425	0.5486	0.1734	44	-3.03	013.08	0
013	Fat, Acid Pretreat (%)	2073	2.470	0.0600	3.425	0.5486	0.1734	44	-1.74	013.00	0
013	Fat, Acid Pretreat (%)	2076	2.491	0.9926	3.425	0.5486	0.1734	44	-1.70	013.00	0
013	Fat, Acid Pretreat (%)	0036	2.539	0.1525	3.425	0.5486	0.1734	44	-1.62	013.10	0
013	Fat, Acid Pretreat (%)	2146	2.850	0.0800	3.425	0.5486	0.1734	44	-1.05	013.00	0
013	Fat, Acid Pretreat (%)	0083	2.865	0.0100	3.425	0.5486	0.1734	44	-1.02	013.00	0
013	Fat, Acid Pretreat (%)	0683	2.915	0.1100	3.425	0.5486	0.1734	44	-0.93	013.13	0
013	Fat, Acid Pretreat (%)	0643	2.920	0.0400	3.425	0.5486	0.1734	44	-0.92	013.02	0
013	Fat, Acid Pretreat (%)	0939	3.000	0.3200	3.425	0.5486	0.1734	44	-0.78	013.10	0
013	Fat, Acid Pretreat (%)	2290	3.100	0.2000	3.425	0.5486	0.1734	44	-0.59	013.12	0
013	Fat, Acid Pretreat (%)	0541	3.125	0.0100	3.425	0.5486	0.1734	44	-0.55	013.00	0
013	Fat, Acid Pretreat (%)	0650	3.145	0.0300	3.425	0.5486	0.1734	44	-0.51	013.00	0
013	Fat, Acid Pretreat (%)	0353	3.170	0.1400	3.425	0.5486	0.1734	44	-0.47	013.10	0
013	Fat, Acid Pretreat (%)	0242	3.185	0.1900	3.425	0.5486	0.1734	44	-0.44	013.00	0
013	Fat, Acid Pretreat (%)	2319	3.200	0.0000	3.425	0.5486	0.1734	44	-0.41	013.00	0
013	Fat, Acid Pretreat (%)	0853	3.200	0.0000	3.425	0.5486	0.1734	44	-0.41	013.02	0
013	Fat, Acid Pretreat (%)	2058	3.245	0.3900	3.425	0.5486	0.1734	44	-0.33	013.02	0
013	Fat, Acid Pretreat (%)	0148	3.255	0.2700	3.425	0.5486	0.1734	44	-0.31	013.02	0
013	Fat, Acid Pretreat (%)	0910	3.260	0.1000	3.425	0.5486	0.1734	44	-0.30	013.00	0
013	Fat, Acid Pretreat (%)	0098	3.280	0.0600	3.425	0.5486	0.1734	44	-0.27	013.02	0
013	Fat, Acid Pretreat (%)	0948	3.290	0.0000	3.425	0.5486	0.1734	44	-0.25	013.00	0
013	Fat, Acid Pretreat (%)	0652	3.300	0.0000	3.425	0.5486	0.1734	44	-0.23	013.00	0
013	Fat, Acid Pretreat (%)	0610	3.325	0.0900	3.425	0.5486	0.1734	44	-0.18	013.10	0
013	Fat, Acid Pretreat (%)	0884	3.350	0.1000	3.425	0.5486	0.1734	44	-0.14	013.00	0
013	Fat, Acid Pretreat (%)	0309	3.360	0.1572	3.425	0.5486	0.1734	44	-0.12	013.00	0
013	Fat, Acid Pretreat (%)	0689	3.450	0.1000	3.425	0.5486	0.1734	44	0.04	013.00	0
013	Fat, Acid Pretreat (%)	0914	3.540	0.1600	3.425	0.5486	0.1734	44	0.21	013.02	0
013	Fat, Acid Pretreat (%)	2196	3.600	0.0000	3.425	0.5486	0.1734	44	0.32	013.10	0
013	Fat, Acid Pretreat (%)	0553	3.635	0.1100	3.425	0.5486	0.1734	44	0.38	013.02	0
013	Fat, Acid Pretreat (%)	0870	3.642	0.0794	3.425	0.5486	0.1734	44	0.40	013.02	0
013	Fat, Acid Pretreat (%)	0208	3.730	0.0600	3.425	0.5486	0.1734	44	0.56	013.02	0
013	Fat, Acid Pretreat (%)	0265	3.780	0.1400	3.425	0.5486	0.1734	44	0.65	013.13	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
013	Fat, Acid Pretreat (%)	0921	3.805	0.1300	3.425	0.5486	0.1734	44	0.69	013.02	0
013	Fat, Acid Pretreat (%)	0682	3.820	0.0000	3.425	0.5486	0.1734	44	0.72	013.02	0
013	Fat, Acid Pretreat (%)	2150	3.925	0.1100	3.425	0.5486	0.1734	44	0.91	013.00	0
013	Fat, Acid Pretreat (%)	0164	3.930	0.0400	3.425	0.5486	0.1734	44	0.92	013.02	0
013	Fat, Acid Pretreat (%)	2328	3.940	0.1000	3.425	0.5486	0.1734	44	0.94	013.02	0
013	Fat, Acid Pretreat (%)	0202	3.965	0.1900	3.425	0.5486	0.1734	44	0.98	013.00	0
013	Fat, Acid Pretreat (%)	0581	4.075	0.0900	3.425	0.5486	0.1734	44	1.18	013.13	0
013	Fat, Acid Pretreat (%)	0407	4.170	0.1170	3.425	0.5486	0.1734	44	1.36	013.02	0
013	Fat, Acid Pretreat (%)	0618	4.218	0.3908	3.425	0.5486	0.1734	44	1.44	013.00	0
013	Fat, Acid Pretreat (%)	0171	4.270	0.1400	3.425	0.5486	0.1734	44	1.54	013.02	0
013	Fat, Acid Pretreat (%)	0843	4.450	1.020	3.425	0.5486	0.1734	44	1.87	013.13	0
013	Fat, Acid Pretreat (%)	0407	4.551	0.6672	3.425	0.5486	0.1734	44	2.05	013.13	0
013	Fat, Acid Pretreat (%)	0074	4.785	2.030	3.425	0.5486	0.1734	44	2.48	013.13	1
015	Aluminum (ppm)	0425	54.75	8.700	85.64	18.92	5.760	19	-1.63	015.43	0
015	Aluminum (ppm)	0037	59.35	0.5000	85.64	18.92	5.760	19	-1.39	015.42	0
015	Aluminum (ppm)	0918	60.20	10.80	85.64	18.92	5.760	19	-1.34	015.43	0
015	Aluminum (ppm)	0049	67.42	9.310	85.64	18.92	5.760	19	-0.96	015.41	0
015	Aluminum (ppm)	0510	73.50	1.000	85.64	18.92	5.760	19	-0.64	015.43	0
015	Aluminum (ppm)	2129	75.98	13.55	85.64	18.92	5.760	19	-0.51	015.42	0
015	Aluminum (ppm)	0297	79.50	1.000	85.64	18.92	5.760	19	-0.32	015.43	0
015	Aluminum (ppm)	0560	81.80	16.00	85.64	18.92	5.760	19	-0.20	015.52	0
015	Aluminum (ppm)	0407	83.70	0.6301	85.64	18.92	5.760	19	-0.10	015.41	0
015	Aluminum (ppm)	0171	88.94	0.8800	85.64	18.92	5.760	19	0.17	015.41	0
015	Aluminum (ppm)	0164	94.00	2.000	85.64	18.92	5.760	19	0.44	015.41	0
015	Aluminum (ppm)	0345	94.05	3.900	85.64	18.92	5.760	19	0.44	015.43	0
015	Aluminum (ppm)	0553	94.30	4.000	85.64	18.92	5.760	19	0.46	015.53	0
015	Aluminum (ppm)	0520	96.50	3.000	85.64	18.92	5.760	19	0.57	015.41	0
015	Aluminum (ppm)	0407	96.73	2.003	85.64	18.92	5.760	19	0.59	015.43	0
015	Aluminum (ppm)	0021	99.00	2.000	85.64	18.92	5.760	19	0.71	015.43	0
015	Aluminum (ppm)	0353	105.0	6.000	85.64	18.92	5.760	19	1.02	015.43	0
015	Aluminum (ppm)	0939	105.9	12.17	85.64	18.92	5.760	19	1.07	015.53	0
015	Aluminum (ppm)	0169	115.0	12.00	85.64	18.92	5.760	19	1.55	015.43	0
017	Boron (ppm)	0083	10.10	0.4000	11.48	1.004	0.4385	19	-1.38	017.43	0
017	Boron (ppm)	0229	10.32	0.7100	11.48	1.004	0.4385	19	-1.17	017.41	0
017	Boron (ppm)	0171	10.32	0.0200	11.48	1.004	0.4385	19	-1.16	017.41	0
017	Boron (ppm)	0096	10.75	0.5586	11.48	1.004	0.4385	19	-0.74	017.52	0
017	Boron (ppm)	0265	10.80	0.2000	11.48	1.004	0.4385	19	-0.68	017.42	0
017	Boron (ppm)	0345	10.99	0.2200	11.48	1.004	0.4385	19	-0.49	017.43	0
017	Boron (ppm)	0693	11.11	0.3270	11.48	1.004	0.4385	19	-0.37	017.42	0
017	Boron (ppm)	0407	11.12	0.2395	11.48	1.004	0.4385	19	-0.36	017.43	0
017	Boron (ppm)	0045	11.15	0.1000	11.48	1.004	0.4385	19	-0.33	017.42	0
017	Boron (ppm)	0407	11.38	0.2462	11.48	1.004	0.4385	19	-0.10	017.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
017	Boron (ppm)	2129	11.47	0.4300	11.48	1.004	0.4385	19	-0.02	017.42	0
017	Boron (ppm)	0226	11.54	0.2300	11.48	1.004	0.4385	19	0.05	017.41	0
017	Boron (ppm)	0918	11.81	0.0600	11.48	1.004	0.4385	19	0.32	017.43	0
017	Boron (ppm)	0510	12.00	0.0000	11.48	1.004	0.4385	19	0.51	017.43	0
017	Boron (ppm)	0294	12.16	0.1200	11.48	1.004	0.4385	19	0.67	017.42	0
017	Boron (ppm)	0358	12.62	1.790	11.48	1.004	0.4385	19	1.13	017.41	0
017	Boron (ppm)	0037	12.75	0.5000	11.48	1.004	0.4385	19	1.26	017.42	0
017	Boron (ppm)	0049	12.84	0.9000	11.48	1.004	0.4385	19	1.35	017.41	0
017	Boron (ppm)	0560	15.61	1.280	11.48	1.004	0.4385	19	4.11	017.52	0
017	Boron (ppm)	0353	14.00	4.000	11.48	1.004	0.4385	19	2.51	017.43	1
017	Boron (ppm)	0297	0.0000	0.0000	11.48	1.004	0.4385	19		017.43	4
017	Boron (ppm)	0021	< 25		11.48	1.004	0.4385	19		017.43	5
019	Calcium (%)	0018	0.8000	0.0140	1.174	0.0625	0.0274	129	-5.98	019.41	0
019	Calcium (%)	2200	0.8950	0.1300	1.174	0.0625	0.0274	129	-4.46	019.08	0
019	Calcium (%)	2129	0.9610	0.0220	1.174	0.0625	0.0274	129	-3.40	019.42	0
019	Calcium (%)	0175	0.9850	0.0100	1.174	0.0625	0.0274	129	-3.02	019.31	0
019	Calcium (%)	2199	0.9950	0.0500	1.174	0.0625	0.0274	129	-2.86	019.08	0
019	Calcium (%)	0154	1.022	0.0036	1.174	0.0625	0.0274	129	-2.42	019.52	0
019	Calcium (%)	0242	1.045	0.0100	1.174	0.0625	0.0274	129	-2.06	019.99	0
019	Calcium (%)	0047	1.065	0.0100	1.174	0.0625	0.0274	129	-1.74	019.52	0
019	Calcium (%)	2181	1.075	0.0500	1.174	0.0625	0.0274	129	-1.58	019.00	0
019	Calcium (%)	0910	1.080	0.0800	1.174	0.0625	0.0274	129	-1.50	019.41	0
019	Calcium (%)	0939	1.080	0.0600	1.174	0.0625	0.0274	129	-1.50	019.53	0
019	Calcium (%)	0889	1.085	0.0200	1.174	0.0625	0.0274	129	-1.42	019.99	0
019	Calcium (%)	0208	1.092	0.0100	1.174	0.0625	0.0274	129	-1.31	019.41	0
019	Calcium (%)	2006	1.094	0.0500	1.174	0.0625	0.0274	129	-1.27	019.00	0
019	Calcium (%)	0096	1.099	0.0231	1.174	0.0625	0.0274	129	-1.20	019.42	0
019	Calcium (%)	0939	1.100	0.0200	1.174	0.0625	0.0274	129	-1.18	019.31	0
019	Calcium (%)	0425	1.105	0.0500	1.174	0.0625	0.0274	129	-1.10	019.43	0
019	Calcium (%)	0036	1.108	0.0195	1.174	0.0625	0.0274	129	-1.05	019.43	0
019	Calcium (%)	2295	1.110	0.0200	1.174	0.0625	0.0274	129	-1.02	019.00	0
019	Calcium (%)	0263	1.112	0.0068	1.174	0.0625	0.0274	129	-0.99	019.42	0
019	Calcium (%)	0884	1.115	0.0100	1.174	0.0625	0.0274	129	-0.94	019.31	0
019	Calcium (%)	2089	1.115	0.0100	1.174	0.0625	0.0274	129	-0.94	019.43	0
019	Calcium (%)	2246	1.120	0.0000	1.174	0.0625	0.0274	129	-0.86	019.00	0
019	Calcium (%)	2319	1.120	0.0200	1.174	0.0625	0.0274	129	-0.86	019.42	0
019	Calcium (%)	0510	1.120	0.0000	1.174	0.0625	0.0274	129	-0.86	019.43	0
019	Calcium (%)	0357	1.122	0.0110	1.174	0.0625	0.0274	129	-0.83	019.42	0
019	Calcium (%)	0878	1.125	0.0500	1.174	0.0625	0.0274	129	-0.78	019.00	0
019	Calcium (%)	0529	1.125	0.0500	1.174	0.0625	0.0274	129	-0.78	019.31	0
019	Calcium (%)	0019	1.125	0.0100	1.174	0.0625	0.0274	129	-0.78	019.41	0
019	Calcium (%)	0520	1.125	0.0100	1.174	0.0625	0.0274	129	-0.78	019.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
019	Calcium (%)	0590	1.125	0.0100	1.174	0.0625	0.0274	129	-0.78	019.99	0
019	Calcium (%)	0144	1.130	0.0200	1.174	0.0625	0.0274	129	-0.70	019.41	0
019	Calcium (%)	0590	1.135	0.0100	1.174	0.0625	0.0274	129	-0.62	019.08	0
019	Calcium (%)	0164	1.135	0.0100	1.174	0.0625	0.0274	129	-0.62	019.44	0
019	Calcium (%)	0142	1.139	0.0492	1.174	0.0625	0.0274	129	-0.55	019.31	0
019	Calcium (%)	0083	1.140	0.0200	1.174	0.0625	0.0274	129	-0.54	019.43	0
019	Calcium (%)	0297	1.140	0.0000	1.174	0.0625	0.0274	129	-0.54	019.43	0
019	Calcium (%)	0610	1.140	0.0200	1.174	0.0625	0.0274	129	-0.54	019.43	0
019	Calcium (%)	0968	1.142	0.0010	1.174	0.0625	0.0274	129	-0.51	019.43	0
019	Calcium (%)	0687	1.145	0.0300	1.174	0.0625	0.0274	129	-0.46	019.31	0
019	Calcium (%)	0004	1.145	0.0500	1.174	0.0625	0.0274	129	-0.46	019.41	0
019	Calcium (%)	2022	1.150	0.0200	1.174	0.0625	0.0274	129	-0.38	019.00	0
019	Calcium (%)	0171	1.150	0.0000	1.174	0.0625	0.0274	129	-0.38	019.41	0
019	Calcium (%)	0045	1.150	0.0800	1.174	0.0625	0.0274	129	-0.38	019.42	0
019	Calcium (%)	0021	1.150	0.1000	1.174	0.0625	0.0274	129	-0.38	019.43	0
019	Calcium (%)	0089	1.150	0.0000	1.174	0.0625	0.0274	129	-0.38	019.43	0
019	Calcium (%)	0186	1.150	0.0000	1.174	0.0625	0.0274	129	-0.38	019.52	0
019	Calcium (%)	0096	1.151	0.0714	1.174	0.0625	0.0274	129	-0.36	019.52	0
019	Calcium (%)	0893	1.155	0.0500	1.174	0.0625	0.0274	129	-0.30	019.00	0
019	Calcium (%)	2144	1.155	0.0100	1.174	0.0625	0.0274	129	-0.30	019.00	0
019	Calcium (%)	0208	1.156	0.0470	1.174	0.0625	0.0274	129	-0.29	019.31	0
019	Calcium (%)	2246	1.160	0.0000	1.174	0.0625	0.0274	129	-0.22	019.31	0
019	Calcium (%)	0298	1.160	0.0200	1.174	0.0625	0.0274	129	-0.22	019.41	0
019	Calcium (%)	0033	1.160	0.0400	1.174	0.0625	0.0274	129	-0.22	019.43	0
019	Calcium (%)	0300	1.162	0.0370	1.174	0.0625	0.0274	129	-0.19	019.43	0
019	Calcium (%)	0553	1.165	0.0900	1.174	0.0625	0.0274	129	-0.14	019.53	0
019	Calcium (%)	0009	1.168	0.1280	1.174	0.0625	0.0274	129	-0.09	019.42	0
019	Calcium (%)	0074	1.170	0.0000	1.174	0.0625	0.0274	129	-0.06	019.41	0
019	Calcium (%)	0294	1.170	0.0200	1.174	0.0625	0.0274	129	-0.06	019.42	0
019	Calcium (%)	0345	1.170	0.0600	1.174	0.0625	0.0274	129	-0.06	019.43	0
019	Calcium (%)	2109	1.171	0.0290	1.174	0.0625	0.0274	129	-0.05	019.00	0
019	Calcium (%)	0169	1.175	0.0100	1.174	0.0625	0.0274	129	0.02	019.32	0
019	Calcium (%)	0505	1.175	0.0300	1.174	0.0625	0.0274	129	0.02	019.43	0
019	Calcium (%)	0918	1.175	0.0340	1.174	0.0625	0.0274	129	0.02	019.43	0
019	Calcium (%)	0572	1.175	0.0500	1.174	0.0625	0.0274	129	0.02	019.53	0
019	Calcium (%)	0560	1.175	0.0562	1.174	0.0625	0.0274	129	0.03	019.52	0
019	Calcium (%)	0693	1.178	0.0180	1.174	0.0625	0.0274	129	0.07	019.42	0
019	Calcium (%)	2196	1.180	0.0000	1.174	0.0625	0.0274	129	0.10	019.08	0
019	Calcium (%)	2196	1.180	0.0000	1.174	0.0625	0.0274	129	0.10	019.31	0
019	Calcium (%)	0265	1.180	0.1000	1.174	0.0625	0.0274	129	0.10	019.42	0
019	Calcium (%)	0010	1.180	0.0600	1.174	0.0625	0.0274	129	0.10	019.43	0
019	Calcium (%)	0682	1.180	0.0000	1.174	0.0625	0.0274	129	0.10	019.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
019	Calcium (%)	2009	1.183	0.0016	1.174	0.0625	0.0274	129	0.14	019.08	0
019	Calcium (%)	2062	1.184	0.0048	1.174	0.0625	0.0274	129	0.17	019.00	0
019	Calcium (%)	0148	1.185	0.0300	1.174	0.0625	0.0274	129	0.18	019.41	0
019	Calcium (%)	0229	1.185	0.0500	1.174	0.0625	0.0274	129	0.18	019.41	0
019	Calcium (%)	0035	1.185	0.0100	1.174	0.0625	0.0274	129	0.18	019.42	0
019	Calcium (%)	0035	1.185	0.0100	1.174	0.0625	0.0274	129	0.18	019.43	0
019	Calcium (%)	0881	1.188	0.0260	1.174	0.0625	0.0274	129	0.23	019.00	0
019	Calcium (%)	0027	1.189	0.0090	1.174	0.0625	0.0274	129	0.24	019.43	0
019	Calcium (%)	0563	1.189	0.0254	1.174	0.0625	0.0274	129	0.25	019.31	0
019	Calcium (%)	0656	1.190	0.0000	1.174	0.0625	0.0274	129	0.26	019.31	0
019	Calcium (%)	0358	1.190	0.0200	1.174	0.0625	0.0274	129	0.26	019.41	0
019	Calcium (%)	0036	1.191	0.0035	1.174	0.0625	0.0274	129	0.27	019.03	0
019	Calcium (%)	0066	1.194	0.0240	1.174	0.0625	0.0274	129	0.33	019.41	0
019	Calcium (%)	0885	1.195	0.0430	1.174	0.0625	0.0274	129	0.34	019.08	0
019	Calcium (%)	0536	1.195	0.0100	1.174	0.0625	0.0274	129	0.34	019.31	0
019	Calcium (%)	0878	1.195	0.0100	1.174	0.0625	0.0274	129	0.34	019.41	0
019	Calcium (%)	0897	1.200	0.0600	1.174	0.0625	0.0274	129	0.42	019.00	0
019	Calcium (%)	2192	1.200	0.0400	1.174	0.0625	0.0274	129	0.42	019.43	0
019	Calcium (%)	0049	1.205	0.0700	1.174	0.0625	0.0274	129	0.50	019.41	0
019	Calcium (%)	0098	1.205	0.0100	1.174	0.0625	0.0274	129	0.50	019.41	0
019	Calcium (%)	0870	1.207	0.0240	1.174	0.0625	0.0274	129	0.54	019.43	0
019	Calcium (%)	0689	1.210	0.0200	1.174	0.0625	0.0274	129	0.58	019.08	0
019	Calcium (%)	0948	1.210	0.0000	1.174	0.0625	0.0274	129	0.58	019.31	0
019	Calcium (%)	0123	1.215	0.0300	1.174	0.0625	0.0274	129	0.66	019.41	0
019	Calcium (%)	0278	1.215	0.0500	1.174	0.0625	0.0274	129	0.66	019.42	0
019	Calcium (%)	2143	1.215	0.0100	1.174	0.0625	0.0274	129	0.66	019.43	0
019	Calcium (%)	2234	1.215	0.0500	1.174	0.0625	0.0274	129	0.66	019.43	0
019	Calcium (%)	0037	1.215	0.0120	1.174	0.0625	0.0274	129	0.66	019.42	0
019	Calcium (%)	0870	1.217	0.0170	1.174	0.0625	0.0274	129	0.69	019.42	0
019	Calcium (%)	2006	1.218	0.0410	1.174	0.0625	0.0274	129	0.70	019.09	0
019	Calcium (%)	0964	1.219	0.0090	1.174	0.0625	0.0274	129	0.73	019.43	0
019	Calcium (%)	2146	1.220	0.0200	1.174	0.0625	0.0274	129	0.74	019.31	0
019	Calcium (%)	0683	1.225	0.0100	1.174	0.0625	0.0274	129	0.82	019.00	0
019	Calcium (%)	2326	1.225	0.0300	1.174	0.0625	0.0274	129	0.82	019.43	0
019	Calcium (%)	0598	1.229	0.0040	1.174	0.0625	0.0274	129	0.89	019.41	0
019	Calcium (%)	0868	1.230	0.0200	1.174	0.0625	0.0274	129	0.90	019.99	0
019	Calcium (%)	0868	1.235	0.0300	1.174	0.0625	0.0274	129	0.98	019.31	0
019	Calcium (%)	0098	1.235	0.0100	1.174	0.0625	0.0274	129	0.98	019.43	0
019	Calcium (%)	0353	1.235	0.0100	1.174	0.0625	0.0274	129	0.98	019.43	0
019	Calcium (%)	2188	1.245	0.0300	1.174	0.0625	0.0274	129	1.14	019.08	0
019	Calcium (%)	0685	1.245	0.0100	1.174	0.0625	0.0274	129	1.14	019.41	0
019	Calcium (%)	0202	1.251	0.0190	1.174	0.0625	0.0274	129	1.23	019.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
019	Calcium (%)	0226	1.255	0.0100	1.174	0.0625	0.0274	129	1.30	019.41	0
019	Calcium (%)	0186	1.262	0.0160	1.174	0.0625	0.0274	129	1.42	019.42	0
019	Calcium (%)	0407	1.270	0.0273	1.174	0.0625	0.0274	129	1.54	019.43	0
019	Calcium (%)	0108	1.270	0.1200	1.174	0.0625	0.0274	129	1.54	019.31	0
019	Calcium (%)	0407	1.278	0.0053	1.174	0.0625	0.0274	129	1.67	019.41	0
019	Calcium (%)	0619	1.285	0.0100	1.174	0.0625	0.0274	129	1.78	019.41	0
019	Calcium (%)	0001	1.288	0.0140	1.174	0.0625	0.0274	129	1.83	019.31	0
019	Calcium (%)	2302	1.290	0.0000	1.174	0.0625	0.0274	129	1.86	019.00	0
019	Calcium (%)	2302	1.290	0.0000	1.174	0.0625	0.0274	129	1.86	019.99	0
019	Calcium (%)	0536	1.305	0.0100	1.174	0.0625	0.0274	129	2.10	019.02	0
019	Calcium (%)	2193	1.325	0.0100	1.174	0.0625	0.0274	129	2.42	019.08	0
019	Calcium (%)	0100	1.330	0.0600	1.174	0.0625	0.0274	129	2.50	019.99	0
019	Calcium (%)	0190	1.355	0.0300	1.174	0.0625	0.0274	129	2.90	019.42	0
019	Calcium (%)	0692	1.365	0.0700	1.174	0.0625	0.0274	129	3.06	019.42	0
019	Calcium (%)	0726	1.385	0.0300	1.174	0.0625	0.0274	129	3.38	019.42	0
019	Calcium (%)	0042	1.115	0.1900	1.174	0.0625	0.0274	129	-0.94	019.42	1
019	Calcium (%)	0366	1.255	0.1500	1.174	0.0625	0.0274	129	1.30	019.42	1
021	Cobalt (ppm)	2234	0.2400	0.0200	0.7760	0.3307	0.0534	28	-1.62	021.43	0
021	Cobalt (ppm)	0689	0.3000	0.0000	0.7760	0.3307	0.0534	28	-1.44	021.31	0
021	Cobalt (ppm)	0693	0.4520	0.1040	0.7760	0.3307	0.0534	28	-0.98	021.42	0
021	Cobalt (ppm)	0910	0.4550	0.0100	0.7760	0.3307	0.0534	28	-0.97	021.52	0
021	Cobalt (ppm)	0169	0.4900	0.1200	0.7760	0.3307	0.0534	28	-0.86	021.43	0
021	Cobalt (ppm)	0918	0.5150	0.0380	0.7760	0.3307	0.0534	28	-0.79	021.53	0
021	Cobalt (ppm)	0619	0.5205	0.0370	0.7760	0.3307	0.0534	28	-0.77	021.41	0
021	Cobalt (ppm)	0096	0.5484	0.0015	0.7760	0.3307	0.0534	28	-0.69	021.52	0
021	Cobalt (ppm)	0186	0.5650	0.0100	0.7760	0.3307	0.0534	28	-0.64	021.52	0
021	Cobalt (ppm)	0047	0.5750	0.0700	0.7760	0.3307	0.0534	28	-0.61	021.52	0
021	Cobalt (ppm)	0353	0.5785	0.0000	0.7760	0.3307	0.0534	28	-0.60	021.43	0
021	Cobalt (ppm)	0171	0.6760	0.0060	0.7760	0.3307	0.0534	28	-0.30	021.41	0
021	Cobalt (ppm)	0560	0.7000	0.0600	0.7760	0.3307	0.0534	28	-0.23	021.52	0
021	Cobalt (ppm)	0939	0.7600	0.0200	0.7760	0.3307	0.0534	28	-0.05	021.53	0
021	Cobalt (ppm)	0045	0.7700	0.2800	0.7760	0.3307	0.0534	28	-0.02	021.42	0
021	Cobalt (ppm)	0345	0.7850	0.0300	0.7760	0.3307	0.0534	28	0.03	021.43	0
021	Cobalt (ppm)	0939	0.8600	0.0400	0.7760	0.3307	0.0534	28	0.25	021.31	0
021	Cobalt (ppm)	0964	0.9220	0.0720	0.7760	0.3307	0.0534	28	0.44	021.43	0
021	Cobalt (ppm)	0610	0.9550	0.0100	0.7760	0.3307	0.0534	28	0.54	021.53	0
021	Cobalt (ppm)	2326	0.9900	0.0200	0.7760	0.3307	0.0534	28	0.65	021.43	0
021	Cobalt (ppm)	0164	1.000	0.0000	0.7760	0.3307	0.0534	28	0.68	021.31	0
021	Cobalt (ppm)	0407	1.013	0.1455	0.7760	0.3307	0.0534	28	0.72	021.41	0
021	Cobalt (ppm)	0021	1.025	0.1500	0.7760	0.3307	0.0534	28	0.75	021.53	0
021	Cobalt (ppm)	0407	1.082	0.1348	0.7760	0.3307	0.0534	28	0.93	021.43	0
021	Cobalt (ppm)	0510	1.095	0.0100	0.7760	0.3307	0.0534	28	0.96	021.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
021	Cobalt (ppm)	0682	1.400	0.0000	0.7760	0.3307	0.0534	28	1.89	021.43	0
021	Cobalt (ppm)	0066	1.545	0.0570	0.7760	0.3307	0.0534	28	2.32	021.41	0
021	Cobalt (ppm)	0148	2.565	0.0500	0.7760	0.3307	0.0534	28	5.41	021.41	0
021	Cobalt (ppm)	0553	1.147	0.4460	0.7760	0.3307	0.0534	28	1.12	021.53	1
021	Cobalt (ppm)	0021	< 2		0.7760	0.3307	0.0534	28		021.43	5
022	Copper (ppm)	0536	5.825	0.1500	13.07	1.325	0.8380	78	-5.47	022.31	0
022	Copper (ppm)	2188	8.035	0.0500	13.07	1.325	0.8380	78	-3.80	022.31	0
022	Copper (ppm)	0208	10.65	0.7000	13.07	1.325	0.8380	78	-1.83	022.31	0
022	Copper (ppm)	0939	10.92	0.2000	13.07	1.325	0.8380	78	-1.62	022.53	0
022	Copper (ppm)	0294	11.12	0.4400	13.07	1.325	0.8380	78	-1.47	022.42	0
022	Copper (ppm)	0049	11.41	1.790	13.07	1.325	0.8380	78	-1.26	022.41	0
022	Copper (ppm)	0939	11.43	0.0400	13.07	1.325	0.8380	78	-1.24	022.31	0
022	Copper (ppm)	0229	11.46	2.000	13.07	1.325	0.8380	78	-1.22	022.41	0
022	Copper (ppm)	0520	11.50	1.000	13.07	1.325	0.8380	78	-1.19	022.41	0
022	Copper (ppm)	2129	11.57	0.4900	13.07	1.325	0.8380	78	-1.14	022.42	0
022	Copper (ppm)	0619	11.80	0.8000	13.07	1.325	0.8380	78	-0.96	022.41	0
022	Copper (ppm)	0098	11.88	0.7500	13.07	1.325	0.8380	78	-0.90	022.41	0
022	Copper (ppm)	0265	11.90	0.2000	13.07	1.325	0.8380	78	-0.88	022.42	0
022	Copper (ppm)	0208	11.94	0.0900	13.07	1.325	0.8380	78	-0.86	022.41	0
022	Copper (ppm)	0045	11.95	0.3000	13.07	1.325	0.8380	78	-0.85	022.42	0
022	Copper (ppm)	0074	12.00	0.0000	13.07	1.325	0.8380	78	-0.81	022.41	0
022	Copper (ppm)	0918	12.00	0.4200	13.07	1.325	0.8380	78	-0.81	022.43	0
022	Copper (ppm)	0242	12.00	0.0000	13.07	1.325	0.8380	78	-0.81	022.99	0
022	Copper (ppm)	0553	12.10	1.400	13.07	1.325	0.8380	78	-0.73	022.53	0
022	Copper (ppm)	0263	12.21	0.2900	13.07	1.325	0.8380	78	-0.65	022.42	0
022	Copper (ppm)	0278	12.25	1.300	13.07	1.325	0.8380	78	-0.62	022.42	0
022	Copper (ppm)	0407	12.30	1.073	13.07	1.325	0.8380	78	-0.58	022.43	0
022	Copper (ppm)	0948	12.32	0.4300	13.07	1.325	0.8380	78	-0.57	022.33	0
022	Copper (ppm)	2143	12.41	0.0900	13.07	1.325	0.8380	78	-0.50	022.43	0
022	Copper (ppm)	0037	12.50	0.2000	13.07	1.325	0.8380	78	-0.43	022.42	0
022	Copper (ppm)	0510	12.50	1.000	13.07	1.325	0.8380	78	-0.43	022.43	0
022	Copper (ppm)	0964	12.51	0.7167	13.07	1.325	0.8380	78	-0.42	022.43	0
022	Copper (ppm)	0098	12.63	1.380	13.07	1.325	0.8380	78	-0.33	022.43	0
022	Copper (ppm)	0572	12.65	0.1000	13.07	1.325	0.8380	78	-0.32	022.53	0
022	Copper (ppm)	0345	12.69	0.4200	13.07	1.325	0.8380	78	-0.29	022.43	0
022	Copper (ppm)	0148	12.70	1.800	13.07	1.325	0.8380	78	-0.28	022.41	0
022	Copper (ppm)	0407	12.74	0.1900	13.07	1.325	0.8380	78	-0.25	022.41	0
022	Copper (ppm)	0357	12.75	0.0470	13.07	1.325	0.8380	78	-0.25	022.42	0
022	Copper (ppm)	0033	12.75	0.3000	13.07	1.325	0.8380	78	-0.24	022.43	0
022	Copper (ppm)	0171	12.77	0.1800	13.07	1.325	0.8380	78	-0.23	022.41	0
022	Copper (ppm)	0870	12.91	0.2900	13.07	1.325	0.8380	78	-0.12	022.43	0
022	Copper (ppm)	0910	13.00	2.000	13.07	1.325	0.8380	78	-0.05	022.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
022	Copper (ppm)	0083	13.00	0.0000	13.07	1.325	0.8380	78	-0.05	022.43	0
022	Copper (ppm)	0297	13.00	0.0000	13.07	1.325	0.8380	78	-0.05	022.43	0
022	Copper (ppm)	0968	13.00	0.0000	13.07	1.325	0.8380	78	-0.05	022.43	0
022	Copper (ppm)	0164	13.00	0.0000	13.07	1.325	0.8380	78	-0.05	022.44	0
022	Copper (ppm)	0870	13.03	0.0500	13.07	1.325	0.8380	78	-0.03	022.42	0
022	Copper (ppm)	0726	13.04	0.0000	13.07	1.325	0.8380	78	-0.02	022.42	0
022	Copper (ppm)	0186	13.05	0.1000	13.07	1.325	0.8380	78	-0.02	022.52	0
022	Copper (ppm)	0226	13.22	0.3100	13.07	1.325	0.8380	78	0.11	022.41	0
022	Copper (ppm)	0202	13.25	0.7000	13.07	1.325	0.8380	78	0.14	022.43	0
022	Copper (ppm)	0560	13.35	0.5000	13.07	1.325	0.8380	78	0.21	022.52	0
022	Copper (ppm)	0598	13.38	0.0500	13.07	1.325	0.8380	78	0.23	022.41	0
022	Copper (ppm)	0682	13.40	0.0000	13.07	1.325	0.8380	78	0.25	022.43	0
022	Copper (ppm)	0096	13.42	0.6645	13.07	1.325	0.8380	78	0.26	022.52	0
022	Copper (ppm)	0019	13.43	0.1700	13.07	1.325	0.8380	78	0.27	022.41	0
022	Copper (ppm)	0590	13.45	0.9000	13.07	1.325	0.8380	78	0.29	022.99	0
022	Copper (ppm)	0878	13.50	1.000	13.07	1.325	0.8380	78	0.32	022.41	0
022	Copper (ppm)	0021	13.50	3.000	13.07	1.325	0.8380	78	0.32	022.53	0
022	Copper (ppm)	0100	13.50	1.000	13.07	1.325	0.8380	78	0.32	022.99	0
022	Copper (ppm)	0868	13.51	0.2200	13.07	1.325	0.8380	78	0.33	022.31	0
022	Copper (ppm)	0685	13.69	0.0100	13.07	1.325	0.8380	78	0.46	022.42	0
022	Copper (ppm)	0505	13.71	2.820	13.07	1.325	0.8380	78	0.48	022.43	0
022	Copper (ppm)	0425	13.95	0.1000	13.07	1.325	0.8380	78	0.66	022.43	0
022	Copper (ppm)	0175	14.00	0.0000	13.07	1.325	0.8380	78	0.70	022.31	0
022	Copper (ppm)	2302	14.00	0.0000	13.07	1.325	0.8380	78	0.70	022.99	0
022	Copper (ppm)	0692	14.25	2.700	13.07	1.325	0.8380	78	0.89	022.42	0
022	Copper (ppm)	2319	14.25	0.1000	13.07	1.325	0.8380	78	0.89	022.42	0
022	Copper (ppm)	0693	14.26	1.969	13.07	1.325	0.8380	78	0.90	022.42	0
022	Copper (ppm)	0096	14.29	3.835	13.07	1.325	0.8380	78	0.92	022.42	0
022	Copper (ppm)	2196	14.50	0.0000	13.07	1.325	0.8380	78	1.08	022.31	0
022	Copper (ppm)	0169	14.50	4.600	13.07	1.325	0.8380	78	1.08	022.43	0
022	Copper (ppm)	0610	14.70	3.200	13.07	1.325	0.8380	78	1.23	022.43	0
022	Copper (ppm)	0358	14.95	3.330	13.07	1.325	0.8380	78	1.41	022.41	0
022	Copper (ppm)	0190	15.13	0.1900	13.07	1.325	0.8380	78	1.55	022.42	0
022	Copper (ppm)	0042	15.25	2.700	13.07	1.325	0.8380	78	1.65	022.42	0
022	Copper (ppm)	2326	15.26	0.2500	13.07	1.325	0.8380	78	1.65	022.43	0
022	Copper (ppm)	2192	15.35	1.700	13.07	1.325	0.8380	78	1.72	022.43	0
022	Copper (ppm)	0353	15.50	1.000	13.07	1.325	0.8380	78	1.83	022.43	0
022	Copper (ppm)	0035	17.26	0.9900	13.07	1.325	0.8380	78	3.16	022.43	0
022	Copper (ppm)	0529	17.90	1.400	13.07	1.325	0.8380	78	3.65	022.31	0
022	Copper (ppm)	0035	19.41	0.5800	13.07	1.325	0.8380	78	4.79	022.42	0
022	Copper (ppm)	2089	21.40	2.600	13.07	1.325	0.8380	78	6.29	022.43	0
022	Copper (ppm)	0366	17.00	6.000	13.07	1.325	0.8380	78	2.97	022.42	1

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
022	Copper (ppm)	0021	< 25		13.07	1.325	0.8380	78		022.43	5
025	Iron (ppm)	0536	165.1	0.7000	202.9	17.40	6.129	71	-2.17	025.31	0
025	Iron (ppm)	0560	165.6	27.40	202.9	17.40	6.129	71	-2.14	025.52	0
025	Iron (ppm)	0278	166.5	13.00	202.9	17.40	6.129	71	-2.09	025.42	0
025	Iron (ppm)	0294	169.6	0.1800	202.9	17.40	6.129	71	-1.91	025.42	0
025	Iron (ppm)	0265	172.5	11.00	202.9	17.40	6.129	71	-1.75	025.42	0
025	Iron (ppm)	0619	173.3	9.700	202.9	17.40	6.129	71	-1.70	025.41	0
025	Iron (ppm)	0910	176.0	12.00	202.9	17.40	6.129	71	-1.55	025.41	0
025	Iron (ppm)	0726	176.7	0.1000	202.9	17.40	6.129	71	-1.51	025.42	0
025	Iron (ppm)	0226	179.2	11.93	202.9	17.40	6.129	71	-1.36	025.41	0
025	Iron (ppm)	0042	179.5	33.00	202.9	17.40	6.129	71	-1.35	025.42	0
025	Iron (ppm)	2319	182.5	7.000	202.9	17.40	6.129	71	-1.17	025.42	0
025	Iron (ppm)	0045	183.5	7.000	202.9	17.40	6.129	71	-1.12	025.42	0
025	Iron (ppm)	0870	186.0	0.3000	202.9	17.40	6.129	71	-0.98	025.42	0
025	Iron (ppm)	2188	188.5	2.240	202.9	17.40	6.129	71	-0.83	025.31	0
025	Iron (ppm)	0263	190.3	3.380	202.9	17.40	6.129	71	-0.73	025.42	0
025	Iron (ppm)	0096	191.4	7.048	202.9	17.40	6.129	71	-0.66	025.52	0
025	Iron (ppm)	0939	191.9	2.870	202.9	17.40	6.129	71	-0.63	025.31	0
025	Iron (ppm)	0520	192.5	5.000	202.9	17.40	6.129	71	-0.60	025.41	0
025	Iron (ppm)	0345	193.3	1.500	202.9	17.40	6.129	71	-0.56	025.43	0
025	Iron (ppm)	0208	193.5	12.30	202.9	17.40	6.129	71	-0.54	025.41	0
025	Iron (ppm)	0939	193.5	5.000	202.9	17.40	6.129	71	-0.54	025.53	0
025	Iron (ppm)	0242	194.0	10.00	202.9	17.40	6.129	71	-0.51	025.99	0
025	Iron (ppm)	2129	194.2	11.50	202.9	17.40	6.129	71	-0.50	025.42	0
025	Iron (ppm)	0123	195.5	7.000	202.9	17.40	6.129	71	-0.43	025.41	0
025	Iron (ppm)	0164	196.0	0.0000	202.9	17.40	6.129	71	-0.40	025.41	0
025	Iron (ppm)	0353	196.0	6.000	202.9	17.40	6.129	71	-0.40	025.43	0
025	Iron (ppm)	0553	197.5	17.00	202.9	17.40	6.129	71	-0.31	025.53	0
025	Iron (ppm)	0083	198.0	4.000	202.9	17.40	6.129	71	-0.28	025.43	0
025	Iron (ppm)	0096	199.9	27.48	202.9	17.40	6.129	71	-0.17	025.42	0
025	Iron (ppm)	0297	201.0	4.000	202.9	17.40	6.129	71	-0.11	025.43	0
025	Iron (ppm)	0510	201.5	1.000	202.9	17.40	6.129	71	-0.08	025.43	0
025	Iron (ppm)	0033	203.0	6.000	202.9	17.40	6.129	71	0.00	025.43	0
025	Iron (ppm)	0918	203.4	0.2600	202.9	17.40	6.129	71	0.03	025.43	0
025	Iron (ppm)	0074	205.0	0.0000	202.9	17.40	6.129	71	0.12	025.41	0
025	Iron (ppm)	0964	205.0	0.5000	202.9	17.40	6.129	71	0.12	025.43	0
025	Iron (ppm)	0019	205.5	0.5000	202.9	17.40	6.129	71	0.15	025.41	0
025	Iron (ppm)	0868	206.0	2.000	202.9	17.40	6.129	71	0.18	025.31	0
025	Iron (ppm)	0358	206.1	0.5200	202.9	17.40	6.129	71	0.18	025.41	0
025	Iron (ppm)	0878	208.0	2.000	202.9	17.40	6.129	71	0.29	025.41	0
025	Iron (ppm)	0407	208.4	1.087	202.9	17.40	6.129	71	0.32	025.41	0
025	Iron (ppm)	0171	208.5	1.800	202.9	17.40	6.129	71	0.32	025.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
025	Iron (ppm)	0169	208.5	7.000	202.9	17.40	6.129	71	0.32	025.43	0
025	Iron (ppm)	0610	208.5	3.000	202.9	17.40	6.129	71	0.32	025.43	0
025	Iron (ppm)	0968	208.5	1.000	202.9	17.40	6.129	71	0.32	025.43	0
025	Iron (ppm)	0098	208.9	7.300	202.9	17.40	6.129	71	0.34	025.41	0
025	Iron (ppm)	0693	209.1	4.420	202.9	17.40	6.129	71	0.35	025.42	0
025	Iron (ppm)	0021	210.0	0.0000	202.9	17.40	6.129	71	0.41	025.43	0
025	Iron (ppm)	0407	210.1	4.582	202.9	17.40	6.129	71	0.41	025.43	0
025	Iron (ppm)	0190	210.5	1.010	202.9	17.40	6.129	71	0.44	025.42	0
025	Iron (ppm)	2302	211.0	0.0000	202.9	17.40	6.129	71	0.46	025.99	0
025	Iron (ppm)	0870	211.2	5.100	202.9	17.40	6.129	71	0.47	025.43	0
025	Iron (ppm)	0598	214.7	1.100	202.9	17.40	6.129	71	0.67	025.41	0
025	Iron (ppm)	0049	214.7	5.130	202.9	17.40	6.129	71	0.68	025.41	0
025	Iron (ppm)	2196	216.3	0.0000	202.9	17.40	6.129	71	0.77	025.31	0
025	Iron (ppm)	0035	216.9	2.600	202.9	17.40	6.129	71	0.80	025.43	0
025	Iron (ppm)	0208	217.0	16.00	202.9	17.40	6.129	71	0.81	025.31	0
025	Iron (ppm)	0148	217.1	19.79	202.9	17.40	6.129	71	0.82	025.41	0
025	Iron (ppm)	0098	217.2	4.100	202.9	17.40	6.129	71	0.82	025.43	0
025	Iron (ppm)	0175	218.0	0.0000	202.9	17.40	6.129	71	0.87	025.31	0
025	Iron (ppm)	0100	219.0	4.000	202.9	17.40	6.129	71	0.92	025.99	0
025	Iron (ppm)	0035	221.1	3.800	202.9	17.40	6.129	71	1.04	025.42	0
025	Iron (ppm)	0529	221.8	0.5000	202.9	17.40	6.129	71	1.08	025.34	0
025	Iron (ppm)	0505	221.9	24.50	202.9	17.40	6.129	71	1.09	025.43	0
025	Iron (ppm)	2326	222.0	2.000	202.9	17.40	6.129	71	1.10	025.43	0
025	Iron (ppm)	2089	222.3	4.540	202.9	17.40	6.129	71	1.11	025.43	0
025	Iron (ppm)	0037	223.9	0.3000	202.9	17.40	6.129	71	1.20	025.42	0
025	Iron (ppm)	0229	224.1	9.400	202.9	17.40	6.129	71	1.22	025.41	0
025	Iron (ppm)	2192	225.5	3.000	202.9	17.40	6.129	71	1.30	025.43	0
025	Iron (ppm)	0366	227.5	23.00	202.9	17.40	6.129	71	1.41	025.42	0
025	Iron (ppm)	0682	234.5	0.0000	202.9	17.40	6.129	71	1.81	025.43	0
025	Iron (ppm)	0563	241.7	2.693	202.9	17.40	6.129	71	2.23	025.31	0
025	Iron (ppm)	0425	123.1	54.30	202.9	17.40	6.129	71	-4.59	025.43	1
025	Iron (ppm)	0692	226.5	47.00	202.9	17.40	6.129	71	1.35	025.42	1
025	Iron (ppm)	2143	52.29	0.3900	202.9	17.40	6.129	71	-8.66	025.43	2
027	Magnesium (%)	2129	0.1640	0.0000	0.1958	0.0124	0.0044	92	-2.57	027.42	0
027	Magnesium (%)	0889	0.1675	0.0050	0.1958	0.0124	0.0044	92	-2.29	027.99	0
027	Magnesium (%)	0089	0.1700	0.0000	0.1958	0.0124	0.0044	92	-2.09	027.43	0
027	Magnesium (%)	0154	0.1737	0.0002	0.1958	0.0124	0.0044	92	-1.79	027.52	0
027	Magnesium (%)	0910	0.1750	0.0100	0.1958	0.0124	0.0044	92	-1.68	027.41	0
027	Magnesium (%)	0529	0.1755	0.0030	0.1958	0.0124	0.0044	92	-1.64	027.31	0
027	Magnesium (%)	0175	0.1800	0.0000	0.1958	0.0124	0.0044	92	-1.28	027.31	0
027	Magnesium (%)	0242	0.1800	0.0000	0.1958	0.0124	0.0044	92	-1.28	027.99	0
027	Magnesium (%)	0870	0.1830	0.0039	0.1958	0.0124	0.0044	92	-1.04	027.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
027	Magnesium (%)	0870	0.1831	0.0016	0.1958	0.0124	0.0044	92	-1.03	027.42	0
027	Magnesium (%)	0045	0.1835	0.0050	0.1958	0.0124	0.0044	92	-1.00	027.42	0
027	Magnesium (%)	0407	0.1848	0.0002	0.1958	0.0124	0.0044	92	-0.89	027.43	0
027	Magnesium (%)	0868	0.1850	0.0100	0.1958	0.0124	0.0044	92	-0.88	027.31	0
027	Magnesium (%)	0019	0.1850	0.0100	0.1958	0.0124	0.0044	92	-0.88	027.41	0
027	Magnesium (%)	0098	0.1850	0.0100	0.1958	0.0124	0.0044	92	-0.88	027.41	0
027	Magnesium (%)	0229	0.1850	0.0100	0.1958	0.0124	0.0044	92	-0.88	027.41	0
027	Magnesium (%)	0520	0.1850	0.0100	0.1958	0.0124	0.0044	92	-0.88	027.41	0
027	Magnesium (%)	0294	0.1850	0.0100	0.1958	0.0124	0.0044	92	-0.88	027.42	0
027	Magnesium (%)	0021	0.1850	0.0100	0.1958	0.0124	0.0044	92	-0.88	027.43	0
027	Magnesium (%)	2143	0.1850	0.0100	0.1958	0.0124	0.0044	92	-0.88	027.43	0
027	Magnesium (%)	2146	0.1865	0.0030	0.1958	0.0124	0.0044	92	-0.75	027.31	0
027	Magnesium (%)	0096	0.1868	0.0059	0.1958	0.0124	0.0044	92	-0.73	027.42	0
027	Magnesium (%)	0345	0.1870	0.0020	0.1958	0.0124	0.0044	92	-0.71	027.43	0
027	Magnesium (%)	0357	0.1876	0.0036	0.1958	0.0124	0.0044	92	-0.67	027.42	0
027	Magnesium (%)	0407	0.1884	0.0057	0.1958	0.0124	0.0044	92	-0.60	027.41	0
027	Magnesium (%)	2319	0.1885	0.0050	0.1958	0.0124	0.0044	92	-0.59	027.42	0
027	Magnesium (%)	0263	0.1888	0.0004	0.1958	0.0124	0.0044	92	-0.57	027.42	0
027	Magnesium (%)	0939	0.1900	0.0000	0.1958	0.0124	0.0044	92	-0.47	027.31	0
027	Magnesium (%)	0049	0.1900	0.0000	0.1958	0.0124	0.0044	92	-0.47	027.41	0
027	Magnesium (%)	0208	0.1900	0.0060	0.1958	0.0124	0.0044	92	-0.47	027.41	0
027	Magnesium (%)	0265	0.1900	0.0000	0.1958	0.0124	0.0044	92	-0.47	027.42	0
027	Magnesium (%)	0278	0.1900	0.0000	0.1958	0.0124	0.0044	92	-0.47	027.42	0
027	Magnesium (%)	0083	0.1900	0.0000	0.1958	0.0124	0.0044	92	-0.47	027.43	0
027	Magnesium (%)	0425	0.1900	0.0000	0.1958	0.0124	0.0044	92	-0.47	027.43	0
027	Magnesium (%)	0510	0.1900	0.0000	0.1958	0.0124	0.0044	92	-0.47	027.43	0
027	Magnesium (%)	0939	0.1900	0.0000	0.1958	0.0124	0.0044	92	-0.47	027.53	0
027	Magnesium (%)	0096	0.1902	0.0083	0.1958	0.0124	0.0044	92	-0.46	027.52	0
027	Magnesium (%)	0968	0.1905	0.0010	0.1958	0.0124	0.0044	92	-0.43	027.43	0
027	Magnesium (%)	0560	0.1917	0.0025	0.1958	0.0124	0.0044	92	-0.34	027.52	0
027	Magnesium (%)	0033	0.1925	0.0050	0.1958	0.0124	0.0044	92	-0.27	027.43	0
027	Magnesium (%)	0186	0.1926	0.0007	0.1958	0.0124	0.0044	92	-0.26	027.52	0
027	Magnesium (%)	0202	0.1935	0.0010	0.1958	0.0124	0.0044	92	-0.19	027.43	0
027	Magnesium (%)	0572	0.1935	0.0050	0.1958	0.0124	0.0044	92	-0.19	027.53	0
027	Magnesium (%)	0164	0.1940	0.0000	0.1958	0.0124	0.0044	92	-0.15	027.44	0
027	Magnesium (%)	0009	0.1944	0.0143	0.1958	0.0124	0.0044	92	-0.12	027.42	0
027	Magnesium (%)	0004	0.1950	0.0100	0.1958	0.0124	0.0044	92	-0.07	027.41	0
027	Magnesium (%)	0035	0.1950	0.0100	0.1958	0.0124	0.0044	92	-0.07	027.42	0
027	Magnesium (%)	0190	0.1950	0.0100	0.1958	0.0124	0.0044	92	-0.07	027.42	0
027	Magnesium (%)	0353	0.1950	0.0100	0.1958	0.0124	0.0044	92	-0.07	027.43	0
027	Magnesium (%)	2192	0.1950	0.0100	0.1958	0.0124	0.0044	92	-0.07	027.43	0
027	Magnesium (%)	0590	0.1950	0.0100	0.1958	0.0124	0.0044	92	-0.07	027.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
027	Magnesium (%)	0066	0.1960	0.0020	0.1958	0.0124	0.0044	92	0.01	027.41	0
027	Magnesium (%)	0074	0.1975	0.0010	0.1958	0.0124	0.0044	92	0.14	027.41	0
027	Magnesium (%)	0171	0.1980	0.0020	0.1958	0.0124	0.0044	92	0.18	027.41	0
027	Magnesium (%)	0186	0.1994	0.0094	0.1958	0.0124	0.0044	92	0.29	027.42	0
027	Magnesium (%)	0689	0.2000	0.0000	0.1958	0.0124	0.0044	92	0.34	027.31	0
027	Magnesium (%)	2196	0.2000	0.0000	0.1958	0.0124	0.0044	92	0.34	027.31	0
027	Magnesium (%)	0948	0.2000	0.0000	0.1958	0.0124	0.0044	92	0.34	027.33	0
027	Magnesium (%)	0878	0.2000	0.0000	0.1958	0.0124	0.0044	92	0.34	027.41	0
027	Magnesium (%)	0726	0.2000	0.0000	0.1958	0.0124	0.0044	92	0.34	027.42	0
027	Magnesium (%)	0035	0.2000	0.0000	0.1958	0.0124	0.0044	92	0.34	027.43	0
027	Magnesium (%)	0297	0.2000	0.0000	0.1958	0.0124	0.0044	92	0.34	027.43	0
027	Magnesium (%)	0563	0.2007	0.0008	0.1958	0.0124	0.0044	92	0.39	027.31	0
027	Magnesium (%)	0142	0.2008	0.0008	0.1958	0.0124	0.0044	92	0.40	027.31	0
027	Magnesium (%)	0918	0.2010	0.0000	0.1958	0.0124	0.0044	92	0.42	027.43	0
027	Magnesium (%)	0208	0.2015	0.0150	0.1958	0.0124	0.0044	92	0.46	027.31	0
027	Magnesium (%)	0027	0.2030	0.0080	0.1958	0.0124	0.0044	92	0.58	027.43	0
027	Magnesium (%)	0037	0.2040	0.0080	0.1958	0.0124	0.0044	92	0.66	027.42	0
027	Magnesium (%)	0598	0.2041	0.0007	0.1958	0.0124	0.0044	92	0.67	027.41	0
027	Magnesium (%)	0685	0.2050	0.0100	0.1958	0.0124	0.0044	92	0.74	027.42	0
027	Magnesium (%)	0692	0.2050	0.0100	0.1958	0.0124	0.0044	92	0.74	027.42	0
027	Magnesium (%)	0098	0.2050	0.0100	0.1958	0.0124	0.0044	92	0.74	027.43	0
027	Magnesium (%)	2326	0.2060	0.0040	0.1958	0.0124	0.0044	92	0.82	027.43	0
027	Magnesium (%)	0148	0.2065	0.0030	0.1958	0.0124	0.0044	92	0.86	027.41	0
027	Magnesium (%)	0001	0.2070	0.0057	0.1958	0.0124	0.0044	92	0.90	027.31	0
027	Magnesium (%)	0964	0.2076	0.0028	0.1958	0.0124	0.0044	92	0.95	027.43	0
027	Magnesium (%)	0226	0.2100	0.0000	0.1958	0.0124	0.0044	92	1.15	027.41	0
027	Magnesium (%)	0358	0.2100	0.0000	0.1958	0.0124	0.0044	92	1.15	027.41	0
027	Magnesium (%)	0610	0.2100	0.0000	0.1958	0.0124	0.0044	92	1.15	027.43	0
027	Magnesium (%)	0682	0.2100	0.0000	0.1958	0.0124	0.0044	92	1.15	027.43	0
027	Magnesium (%)	2089	0.2100	0.0000	0.1958	0.0124	0.0044	92	1.15	027.43	0
027	Magnesium (%)	0366	0.2100	0.0200	0.1958	0.0124	0.0044	92	1.15	027.42	0
027	Magnesium (%)	0169	0.2150	0.0100	0.1958	0.0124	0.0044	92	1.55	027.32	0
027	Magnesium (%)	2234	0.2150	0.0100	0.1958	0.0124	0.0044	92	1.55	027.43	0
027	Magnesium (%)	0100	0.2150	0.0100	0.1958	0.0124	0.0044	92	1.55	027.99	0
027	Magnesium (%)	0619	0.2175	0.0010	0.1958	0.0124	0.0044	92	1.75	027.41	0
027	Magnesium (%)	0693	0.2200	0.0040	0.1958	0.0124	0.0044	92	1.96	027.42	0
027	Magnesium (%)	2302	0.2200	0.0000	0.1958	0.0124	0.0044	92	1.96	027.99	0
027	Magnesium (%)	0536	0.2250	0.0100	0.1958	0.0124	0.0044	92	2.36	027.31	0
027	Magnesium (%)	2246	0.2300	0.0000	0.1958	0.0124	0.0044	92	2.76	027.31	0
027	Magnesium (%)	0505	0.2300	0.0000	0.1958	0.0124	0.0044	92	2.76	027.43	0
027	Magnesium (%)	0047	0.2350	0.0100	0.1958	0.0124	0.0044	92	3.17	027.52	0
027	Magnesium (%)	0042	0.1870	0.0240	0.1958	0.0124	0.0044	92	-0.71	027.42	1

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
027	Magnesium (%)	0553	0.1930	0.0280	0.1958	0.0124	0.0044	92	-0.23	027.53	1
027	Magnesium (%)	0656	0.3900	0.0000	0.1958	0.0124	0.0044	92	15.71	027.31	2
028	Manganese (ppm)	2234	7.165	0.3500	90.07	6.845	2.925	87	-12.11	028.43	0
028	Manganese (ppm)	0047	7.800	0.0000	90.07	6.845	2.925	87	-12.02	028.52	0
028	Manganese (ppm)	2246	53.59	1.340	90.07	6.845	2.925	87	-5.33	028.34	0
028	Manganese (ppm)	0870	69.86	3.300	90.07	6.845	2.925	87	-2.95	028.42	0
028	Manganese (ppm)	0619	70.60	5.800	90.07	6.845	2.925	87	-2.84	028.41	0
028	Manganese (ppm)	0226	71.57	1.480	90.07	6.845	2.925	87	-2.70	028.41	0
028	Manganese (ppm)	0066	75.11	1.383	90.07	6.845	2.925	87	-2.19	028.44	0
028	Manganese (ppm)	0009	79.50	1.000	90.07	6.845	2.925	87	-1.54	028.42	0
028	Manganese (ppm)	0870	79.91	0.1700	90.07	6.845	2.925	87	-1.49	028.43	0
028	Manganese (ppm)	0553	80.15	7.900	90.07	6.845	2.925	87	-1.45	028.53	0
028	Manganese (ppm)	0536	80.17	2.030	90.07	6.845	2.925	87	-1.45	028.31	0
028	Manganese (ppm)	0520	80.50	1.000	90.07	6.845	2.925	87	-1.40	028.41	0
028	Manganese (ppm)	2129	80.79	1.190	90.07	6.845	2.925	87	-1.36	028.42	0
028	Manganese (ppm)	0263	82.87	3.330	90.07	6.845	2.925	87	-1.05	028.42	0
028	Manganese (ppm)	0693	82.95	1.351	90.07	6.845	2.925	87	-1.04	028.42	0
028	Manganese (ppm)	0033	84.15	1.500	90.07	6.845	2.925	87	-0.87	028.43	0
028	Manganese (ppm)	0910	84.50	1.000	90.07	6.845	2.925	87	-0.81	028.41	0
028	Manganese (ppm)	0042	84.75	6.300	90.07	6.845	2.925	87	-0.78	028.42	0
028	Manganese (ppm)	0425	84.90	3.800	90.07	6.845	2.925	87	-0.76	028.43	0
028	Manganese (ppm)	2319	85.55	1.500	90.07	6.845	2.925	87	-0.66	028.42	0
028	Manganese (ppm)	0049	85.81	3.360	90.07	6.845	2.925	87	-0.62	028.41	0
028	Manganese (ppm)	0045	85.85	7.500	90.07	6.845	2.925	87	-0.62	028.42	0
028	Manganese (ppm)	0074	86.00	2.000	90.07	6.845	2.925	87	-0.59	028.41	0
028	Manganese (ppm)	2143	86.49	0.7500	90.07	6.845	2.925	87	-0.52	028.43	0
028	Manganese (ppm)	0692	86.50	7.000	90.07	6.845	2.925	87	-0.52	028.42	0
028	Manganese (ppm)	0505	86.70	2.410	90.07	6.845	2.925	87	-0.49	028.43	0
028	Manganese (ppm)	0685	86.98	0.0100	90.07	6.845	2.925	87	-0.45	028.42	0
028	Manganese (ppm)	0175	87.00	2.000	90.07	6.845	2.925	87	-0.45	028.31	0
028	Manganese (ppm)	0242	87.00	2.000	90.07	6.845	2.925	87	-0.45	028.99	0
028	Manganese (ppm)	0510	87.50	1.000	90.07	6.845	2.925	87	-0.38	028.43	0
028	Manganese (ppm)	0563	87.55	4.829	90.07	6.845	2.925	87	-0.37	028.31	0
028	Manganese (ppm)	0948	87.56	4.710	90.07	6.845	2.925	87	-0.37	028.33	0
028	Manganese (ppm)	0918	87.79	1.390	90.07	6.845	2.925	87	-0.33	028.43	0
028	Manganese (ppm)	0019	87.91	1.040	90.07	6.845	2.925	87	-0.32	028.41	0
028	Manganese (ppm)	0878	88.00	4.000	90.07	6.845	2.925	87	-0.30	028.41	0
028	Manganese (ppm)	0939	88.41	4.600	90.07	6.845	2.925	87	-0.24	028.53	0
028	Manganese (ppm)	0164	88.50	1.000	90.07	6.845	2.925	87	-0.23	028.44	0
028	Manganese (ppm)	0968	89.00	2.000	90.07	6.845	2.925	87	-0.16	028.43	0
028	Manganese (ppm)	0096	89.66	7.765	90.07	6.845	2.925	87	-0.06	028.52	0
028	Manganese (ppm)	2196	89.72	0.0000	90.07	6.845	2.925	87	-0.05	028.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
028	Manganese (ppm)	0939	90.34	1.780	90.07	6.845	2.925	87	0.04	028.31	0
028	Manganese (ppm)	0098	90.38	3.250	90.07	6.845	2.925	87	0.04	028.41	0
028	Manganese (ppm)	0021	91.00	2.000	90.07	6.845	2.925	87	0.14	028.43	0
028	Manganese (ppm)	2302	91.00	0.0000	90.07	6.845	2.925	87	0.14	028.99	0
028	Manganese (ppm)	0297	91.50	3.000	90.07	6.845	2.925	87	0.21	028.43	0
028	Manganese (ppm)	0171	91.53	0.5000	90.07	6.845	2.925	87	0.21	028.41	0
028	Manganese (ppm)	2192	91.65	12.70	90.07	6.845	2.925	87	0.23	028.53	0
028	Manganese (ppm)	0345	91.89	4.620	90.07	6.845	2.925	87	0.27	028.43	0
028	Manganese (ppm)	0610	92.60	0.6000	90.07	6.845	2.925	87	0.37	028.43	0
028	Manganese (ppm)	0407	92.63	2.779	90.07	6.845	2.925	87	0.37	028.43	0
028	Manganese (ppm)	0278	93.00	4.000	90.07	6.845	2.925	87	0.43	028.42	0
028	Manganese (ppm)	0656	93.19	0.0000	90.07	6.845	2.925	87	0.46	028.31	0
028	Manganese (ppm)	0229	93.36	8.900	90.07	6.845	2.925	87	0.48	028.41	0
028	Manganese (ppm)	0358	93.47	3.590	90.07	6.845	2.925	87	0.50	028.41	0
028	Manganese (ppm)	0265	93.50	7.400	90.07	6.845	2.925	87	0.50	028.42	0
028	Manganese (ppm)	0407	93.62	0.1222	90.07	6.845	2.925	87	0.52	028.41	0
028	Manganese (ppm)	0098	93.70	5.020	90.07	6.845	2.925	87	0.53	028.43	0
028	Manganese (ppm)	0169	93.85	1.100	90.07	6.845	2.925	87	0.55	028.43	0
028	Manganese (ppm)	0004	93.87	1.420	90.07	6.845	2.925	87	0.55	028.41	0
028	Manganese (ppm)	0590	93.95	7.100	90.07	6.845	2.925	87	0.57	028.99	0
028	Manganese (ppm)	0353	94.00	2.000	90.07	6.845	2.925	87	0.57	028.43	0
028	Manganese (ppm)	0294	94.10	1.600	90.07	6.845	2.925	87	0.59	028.42	0
028	Manganese (ppm)	0868	94.15	0.9000	90.07	6.845	2.925	87	0.60	028.31	0
028	Manganese (ppm)	0027	94.54	4.543	90.07	6.845	2.925	87	0.65	028.43	0
028	Manganese (ppm)	0357	94.67	1.203	90.07	6.845	2.925	87	0.67	028.42	0
028	Manganese (ppm)	0572	94.75	1.500	90.07	6.845	2.925	87	0.68	028.53	0
028	Manganese (ppm)	0529	94.85	1.100	90.07	6.845	2.925	87	0.70	028.31	0
028	Manganese (ppm)	2188	94.90	0.1300	90.07	6.845	2.925	87	0.70	028.31	0
028	Manganese (ppm)	0208	94.90	2.400	90.07	6.845	2.925	87	0.71	028.31	0
028	Manganese (ppm)	0964	95.02	9.433	90.07	6.845	2.925	87	0.72	028.43	0
028	Manganese (ppm)	0186	95.35	4.500	90.07	6.845	2.925	87	0.77	028.42	0
028	Manganese (ppm)	0083	95.50	1.000	90.07	6.845	2.925	87	0.79	028.43	0
028	Manganese (ppm)	0035	95.67	2.180	90.07	6.845	2.925	87	0.82	028.42	0
028	Manganese (ppm)	0560	95.75	10.30	90.07	6.845	2.925	87	0.83	028.52	0
028	Manganese (ppm)	0148	95.82	0.2600	90.07	6.845	2.925	87	0.84	028.41	0
028	Manganese (ppm)	0598	96.09	0.3900	90.07	6.845	2.925	87	0.88	028.41	0
028	Manganese (ppm)	0186	96.20	1.200	90.07	6.845	2.925	87	0.90	028.52	0
028	Manganese (ppm)	0037	96.35	10.30	90.07	6.845	2.925	87	0.92	028.42	0
028	Manganese (ppm)	0366	97.00	8.000	90.07	6.845	2.925	87	1.01	028.42	0
028	Manganese (ppm)	0202	97.70	4.600	90.07	6.845	2.925	87	1.11	028.43	0
028	Manganese (ppm)	0035	97.88	2.410	90.07	6.845	2.925	87	1.14	028.43	0
028	Manganese (ppm)	2089	99.91	1.920	90.07	6.845	2.925	87	1.44	028.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
028	Manganese (ppm)	0682	100.6	0.0000	90.07	6.845	2.925	87	1.53	028.43	0
028	Manganese (ppm)	2326	101.5	3.000	90.07	6.845	2.925	87	1.67	028.43	0
028	Manganese (ppm)	0190	102.2	1.600	90.07	6.845	2.925	87	1.77	028.42	0
028	Manganese (ppm)	0100	106.5	3.000	90.07	6.845	2.925	87	2.40	028.99	0
028	Manganese (ppm)	0726	107.3	0.0000	90.07	6.845	2.925	87	2.52	028.42	0
028	Manganese (ppm)	0096	89.82	16.13	90.07	6.845	2.925	87	-0.04	028.42	1
031	Phosphorus (%)	0889	0.5325	0.0150	0.6758	0.0330	0.0151	125	-4.34	031.99	0
031	Phosphorus (%)	0108	0.5600	0.0600	0.6758	0.0330	0.0151	125	-3.51	031.01	0
031	Phosphorus (%)	0089	0.5600	0.0000	0.6758	0.0330	0.0151	125	-3.51	031.43	0
031	Phosphorus (%)	2302	0.5800	0.0000	0.6758	0.0330	0.0151	125	-2.91	031.01	0
031	Phosphorus (%)	2302	0.5800	0.0000	0.6758	0.0330	0.0151	125	-2.91	031.99	0
031	Phosphorus (%)	0693	0.6010	0.0400	0.6758	0.0330	0.0151	125	-2.27	031.42	0
031	Phosphorus (%)	2144	0.6050	0.0100	0.6758	0.0330	0.0151	125	-2.15	031.01	0
031	Phosphorus (%)	2146	0.6050	0.0100	0.6758	0.0330	0.0151	125	-2.15	031.01	0
031	Phosphorus (%)	0425	0.6100	0.0200	0.6758	0.0330	0.0151	125	-2.00	031.43	0
031	Phosphorus (%)	2290	0.6239	0.0021	0.6758	0.0330	0.0151	125	-1.58	031.99	0
031	Phosphorus (%)	0278	0.6250	0.0300	0.6758	0.0330	0.0151	125	-1.54	031.42	0
031	Phosphorus (%)	0689	0.6300	0.0200	0.6758	0.0330	0.0151	125	-1.39	031.01	0
031	Phosphorus (%)	2129	0.6335	0.0370	0.6758	0.0330	0.0151	125	-1.28	031.42	0
031	Phosphorus (%)	0175	0.6350	0.0100	0.6758	0.0330	0.0151	125	-1.24	031.01	0
031	Phosphorus (%)	0939	0.6350	0.0100	0.6758	0.0330	0.0151	125	-1.24	031.01	0
031	Phosphorus (%)	2006	0.6365	0.0070	0.6758	0.0330	0.0151	125	-1.19	031.01	0
031	Phosphorus (%)	0019	0.6400	0.0200	0.6758	0.0330	0.0151	125	-1.09	031.41	0
031	Phosphorus (%)	0047	0.6400	0.0200	0.6758	0.0330	0.0151	125	-1.09	031.42	0
031	Phosphorus (%)	0297	0.6400	0.0000	0.6758	0.0330	0.0151	125	-1.09	031.43	0
031	Phosphorus (%)	0242	0.6400	0.0200	0.6758	0.0330	0.0151	125	-1.09	031.99	0
031	Phosphorus (%)	0045	0.6430	0.0500	0.6758	0.0330	0.0151	125	-1.00	031.42	0
031	Phosphorus (%)	0626	0.6435	0.0250	0.6758	0.0330	0.0151	125	-0.98	031.01	0
031	Phosphorus (%)	0169	0.6450	0.0100	0.6758	0.0330	0.0151	125	-0.93	031.03	0
031	Phosphorus (%)	2319	0.6460	0.0260	0.6758	0.0330	0.0151	125	-0.90	031.42	0
031	Phosphorus (%)	0096	0.6471	0.0070	0.6758	0.0330	0.0151	125	-0.87	031.42	0
031	Phosphorus (%)	0897	0.6500	0.0000	0.6758	0.0330	0.0151	125	-0.78	031.01	0
031	Phosphorus (%)	0520	0.6500	0.0000	0.6758	0.0330	0.0151	125	-0.78	031.41	0
031	Phosphorus (%)	0910	0.6500	0.0600	0.6758	0.0330	0.0151	125	-0.78	031.41	0
031	Phosphorus (%)	0294	0.6500	0.0000	0.6758	0.0330	0.0151	125	-0.78	031.42	0
031	Phosphorus (%)	0345	0.6500	0.0200	0.6758	0.0330	0.0151	125	-0.78	031.43	0
031	Phosphorus (%)	0939	0.6500	0.0000	0.6758	0.0330	0.0151	125	-0.78	031.53	0
031	Phosphorus (%)	0884	0.6550	0.0500	0.6758	0.0330	0.0151	125	-0.63	031.01	0
031	Phosphorus (%)	0083	0.6550	0.0100	0.6758	0.0330	0.0151	125	-0.63	031.43	0
031	Phosphorus (%)	0009	0.6564	0.0303	0.6758	0.0330	0.0151	125	-0.59	031.42	0
031	Phosphorus (%)	0208	0.6565	0.0390	0.6758	0.0330	0.0151	125	-0.59	031.41	0
031	Phosphorus (%)	0001	0.6570	0.0020	0.6758	0.0330	0.0151	125	-0.57	031.03	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	0164	0.6585	0.0010	0.6758	0.0330	0.0151	125	-0.53	031.44	0
031	Phosphorus (%)	0298	0.6600	0.0200	0.6758	0.0330	0.0151	125	-0.48	031.41	0
031	Phosphorus (%)	0021	0.6600	0.0200	0.6758	0.0330	0.0151	125	-0.48	031.43	0
031	Phosphorus (%)	2188	0.6650	0.0100	0.6758	0.0330	0.0151	125	-0.33	031.01	0
031	Phosphorus (%)	2200	0.6650	0.0300	0.6758	0.0330	0.0151	125	-0.33	031.01	0
031	Phosphorus (%)	2246	0.6650	0.0100	0.6758	0.0330	0.0151	125	-0.33	031.01	0
031	Phosphorus (%)	0018	0.6655	0.0090	0.6758	0.0330	0.0151	125	-0.31	031.01	0
031	Phosphorus (%)	0553	0.6655	0.0910	0.6758	0.0330	0.0151	125	-0.31	031.53	0
031	Phosphorus (%)	0560	0.6663	0.0265	0.6758	0.0330	0.0151	125	-0.29	031.52	0
031	Phosphorus (%)	0263	0.6679	0.0046	0.6758	0.0330	0.0151	125	-0.24	031.42	0
031	Phosphorus (%)	2109	0.6680	0.0180	0.6758	0.0330	0.0151	125	-0.24	031.01	0
031	Phosphorus (%)	2192	0.6700	0.0000	0.6758	0.0330	0.0151	125	-0.18	031.01	0
031	Phosphorus (%)	0563	0.6721	0.0027	0.6758	0.0330	0.0151	125	-0.11	031.01	0
031	Phosphorus (%)	2326	0.6725	0.0110	0.6758	0.0330	0.0151	125	-0.10	031.43	0
031	Phosphorus (%)	0098	0.6738	0.0025	0.6758	0.0330	0.0151	125	-0.06	031.41	0
031	Phosphorus (%)	0001	0.6740	0.0320	0.6758	0.0330	0.0151	125	-0.06	031.43	0
031	Phosphorus (%)	2009	0.6743	0.0018	0.6758	0.0330	0.0151	125	-0.05	031.01	0
031	Phosphorus (%)	0687	0.6750	0.0500	0.6758	0.0330	0.0151	125	-0.03	031.01	0
031	Phosphorus (%)	0948	0.6750	0.0100	0.6758	0.0330	0.0151	125	-0.03	031.01	0
031	Phosphorus (%)	2022	0.6750	0.0300	0.6758	0.0330	0.0151	125	-0.03	031.01	0
031	Phosphorus (%)	0685	0.6750	0.0100	0.6758	0.0330	0.0151	125	-0.03	031.41	0
031	Phosphorus (%)	0010	0.6750	0.0100	0.6758	0.0330	0.0151	125	-0.03	031.43	0
031	Phosphorus (%)	0357	0.6764	0.0169	0.6758	0.0330	0.0151	125	0.02	031.42	0
031	Phosphorus (%)	0033	0.6775	0.0050	0.6758	0.0330	0.0151	125	0.05	031.43	0
031	Phosphorus (%)	0154	0.6783	0.0073	0.6758	0.0330	0.0151	125	0.07	031.52	0
031	Phosphorus (%)	0885	0.6790	0.0123	0.6758	0.0330	0.0151	125	0.09	031.01	0
031	Phosphorus (%)	0572	0.6790	0.0340	0.6758	0.0330	0.0151	125	0.10	031.53	0
031	Phosphorus (%)	0300	0.6793	0.0342	0.6758	0.0330	0.0151	125	0.10	031.43	0
031	Phosphorus (%)	0868	0.6800	0.0200	0.6758	0.0330	0.0151	125	0.13	031.01	0
031	Phosphorus (%)	0881	0.6800	0.0160	0.6758	0.0330	0.0151	125	0.13	031.01	0
031	Phosphorus (%)	0229	0.6800	0.0200	0.6758	0.0330	0.0151	125	0.13	031.41	0
031	Phosphorus (%)	0683	0.6800	0.0000	0.6758	0.0330	0.0151	125	0.13	031.01	0
031	Phosphorus (%)	2058	0.6800	0.0000	0.6758	0.0330	0.0151	125	0.13	031.01	0
031	Phosphorus (%)	2196	0.6800	0.0000	0.6758	0.0330	0.0151	125	0.13	031.01	0
031	Phosphorus (%)	0144	0.6800	0.0000	0.6758	0.0330	0.0151	125	0.13	031.41	0
031	Phosphorus (%)	0171	0.6800	0.0000	0.6758	0.0330	0.0151	125	0.13	031.41	0
031	Phosphorus (%)	0265	0.6800	0.0000	0.6758	0.0330	0.0151	125	0.13	031.42	0
031	Phosphorus (%)	0027	0.6815	0.0030	0.6758	0.0330	0.0151	125	0.17	031.43	0
031	Phosphorus (%)	0066	0.6815	0.0230	0.6758	0.0330	0.0151	125	0.17	031.44	0
031	Phosphorus (%)	0186	0.6819	0.0041	0.6758	0.0330	0.0151	125	0.18	031.52	0
031	Phosphorus (%)	0208	0.6825	0.0070	0.6758	0.0330	0.0151	125	0.20	031.01	0
031	Phosphorus (%)	0893	0.6850	0.0100	0.6758	0.0330	0.0151	125	0.28	031.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	2295	0.6850	0.0100	0.6758	0.0330	0.0151	125	0.28	031.00	0
031	Phosphorus (%)	0590	0.6850	0.0100	0.6758	0.0330	0.0151	125	0.28	031.01	0
031	Phosphorus (%)	0004	0.6850	0.0100	0.6758	0.0330	0.0151	125	0.28	031.41	0
031	Phosphorus (%)	0049	0.6850	0.0100	0.6758	0.0330	0.0151	125	0.28	031.41	0
031	Phosphorus (%)	0226	0.6850	0.0100	0.6758	0.0330	0.0151	125	0.28	031.41	0
031	Phosphorus (%)	0035	0.6850	0.0100	0.6758	0.0330	0.0151	125	0.28	031.42	0
031	Phosphorus (%)	0353	0.6850	0.0300	0.6758	0.0330	0.0151	125	0.28	031.43	0
031	Phosphorus (%)	2143	0.6850	0.0100	0.6758	0.0330	0.0151	125	0.28	031.43	0
031	Phosphorus (%)	0870	0.6870	0.0307	0.6758	0.0330	0.0151	125	0.34	031.42	0
031	Phosphorus (%)	0036	0.6874	0.0031	0.6758	0.0330	0.0151	125	0.35	031.43	0
031	Phosphorus (%)	0529	0.6900	0.0000	0.6758	0.0330	0.0151	125	0.43	031.01	0
031	Phosphorus (%)	0728	0.6900	0.0000	0.6758	0.0330	0.0151	125	0.43	031.01	0
031	Phosphorus (%)	0035	0.6900	0.0200	0.6758	0.0330	0.0151	125	0.43	031.43	0
031	Phosphorus (%)	0036	0.6914	0.0136	0.6758	0.0330	0.0151	125	0.47	031.03	0
031	Phosphorus (%)	2062	0.6922	0.0050	0.6758	0.0330	0.0151	125	0.50	031.01	0
031	Phosphorus (%)	0918	0.6940	0.0040	0.6758	0.0330	0.0151	125	0.55	031.43	0
031	Phosphorus (%)	0098	0.6944	0.0438	0.6758	0.0330	0.0151	125	0.56	031.43	0
031	Phosphorus (%)	0505	0.6950	0.0100	0.6758	0.0330	0.0151	125	0.58	031.43	0
031	Phosphorus (%)	0142	0.6968	0.0000	0.6758	0.0330	0.0151	125	0.64	031.01	0
031	Phosphorus (%)	0870	0.6968	0.0248	0.6758	0.0330	0.0151	125	0.64	031.43	0
031	Phosphorus (%)	0878	0.7000	0.0200	0.6758	0.0330	0.0151	125	0.73	031.01	0
031	Phosphorus (%)	0610	0.7000	0.0000	0.6758	0.0330	0.0151	125	0.73	031.43	0
031	Phosphorus (%)	0186	0.7010	0.0178	0.6758	0.0330	0.0151	125	0.76	031.42	0
031	Phosphorus (%)	2181	0.7050	0.0100	0.6758	0.0330	0.0151	125	0.88	031.01	0
031	Phosphorus (%)	0510	0.7050	0.0100	0.6758	0.0330	0.0151	125	0.88	031.43	0
031	Phosphorus (%)	0407	0.7065	0.0025	0.6758	0.0330	0.0151	125	0.93	031.43	0
031	Phosphorus (%)	0037	0.7095	0.0150	0.6758	0.0330	0.0151	125	1.02	031.42	0
031	Phosphorus (%)	0202	0.7095	0.0050	0.6758	0.0330	0.0151	125	1.02	031.43	0
031	Phosphorus (%)	0726	0.7100	0.0000	0.6758	0.0330	0.0151	125	1.04	031.42	0
031	Phosphorus (%)	0968	0.7125	0.0010	0.6758	0.0330	0.0151	125	1.11	031.43	0
031	Phosphorus (%)	2089	0.7150	0.0100	0.6758	0.0330	0.0151	125	1.19	031.43	0
031	Phosphorus (%)	0074	0.7165	0.0110	0.6758	0.0330	0.0151	125	1.23	031.41	0
031	Phosphorus (%)	0964	0.7168	0.0168	0.6758	0.0330	0.0151	125	1.24	031.43	0
031	Phosphorus (%)	0123	0.7200	0.0000	0.6758	0.0330	0.0151	125	1.34	031.41	0
031	Phosphorus (%)	0100	0.7200	0.0000	0.6758	0.0330	0.0151	125	1.34	031.99	0
031	Phosphorus (%)	0190	0.7250	0.0100	0.6758	0.0330	0.0151	125	1.49	031.42	0
031	Phosphorus (%)	0619	0.7275	0.0070	0.6758	0.0330	0.0151	125	1.57	031.41	0
031	Phosphorus (%)	0407	0.7295	0.0117	0.6758	0.0330	0.0151	125	1.62	031.41	0
031	Phosphorus (%)	0682	0.7300	0.0000	0.6758	0.0330	0.0151	125	1.64	031.43	0
031	Phosphorus (%)	2234	0.7300	0.0200	0.6758	0.0330	0.0151	125	1.64	031.43	0
031	Phosphorus (%)	0598	0.7325	0.0023	0.6758	0.0330	0.0151	125	1.72	031.41	0
031	Phosphorus (%)	0366	0.7350	0.0900	0.6758	0.0330	0.0151	125	1.79	031.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	2193	0.7400	0.0000	0.6758	0.0330	0.0151	125	1.94	031.01	0
031	Phosphorus (%)	0358	0.7400	0.0600	0.6758	0.0330	0.0151	125	1.94	031.41	0
031	Phosphorus (%)	0656	0.7500	0.0000	0.6758	0.0330	0.0151	125	2.25	031.01	0
031	Phosphorus (%)	0148	0.7505	0.0130	0.6758	0.0330	0.0151	125	2.26	031.41	0
031	Phosphorus (%)	0536	0.7700	0.0200	0.6758	0.0330	0.0151	125	2.85	031.06	0
031	Phosphorus (%)	0042	0.6895	0.1170	0.6758	0.0330	0.0151	125	0.41	031.42	1
031	Phosphorus (%)	0692	0.7400	0.1600	0.6758	0.0330	0.0151	125	1.94	031.42	1
031	Phosphorus (%)	2199	0.9600	0.0600	0.6758	0.0330	0.0151	125	8.61	031.01	2
032	Potassium (%)	0563	0.8463	0.0032	1.008	0.0499	0.0211	77	-3.25	032.31	0
032	Potassium (%)	0154	0.8525	0.0427	1.008	0.0499	0.0211	77	-3.13	032.52	0
032	Potassium (%)	0019	0.8750	0.0300	1.008	0.0499	0.0211	77	-2.67	032.41	0
032	Potassium (%)	0910	0.9050	0.0500	1.008	0.0499	0.0211	77	-2.07	032.41	0
032	Potassium (%)	2129	0.9100	0.0320	1.008	0.0499	0.0211	77	-1.97	032.42	0
032	Potassium (%)	0208	0.9200	0.0640	1.008	0.0499	0.0211	77	-1.77	032.41	0
032	Potassium (%)	0089	0.9400	0.0000	1.008	0.0499	0.0211	77	-1.37	032.43	0
032	Potassium (%)	0425	0.9400	0.0400	1.008	0.0499	0.0211	77	-1.37	032.43	0
032	Potassium (%)	2302	0.9400	0.0000	1.008	0.0499	0.0211	77	-1.37	032.99	0
032	Potassium (%)	0142	0.9451	0.0083	1.008	0.0499	0.0211	77	-1.27	032.31	0
032	Potassium (%)	0242	0.9500	0.0000	1.008	0.0499	0.0211	77	-1.17	032.99	0
032	Potassium (%)	0536	0.9550	0.0100	1.008	0.0499	0.0211	77	-1.07	032.31	0
032	Potassium (%)	2143	0.9550	0.0100	1.008	0.0499	0.0211	77	-1.07	032.43	0
032	Potassium (%)	0553	0.9555	0.0410	1.008	0.0499	0.0211	77	-1.06	032.53	0
032	Potassium (%)	2089	0.9600	0.0000	1.008	0.0499	0.0211	77	-0.97	032.43	0
032	Potassium (%)	0357	0.9666	0.0018	1.008	0.0499	0.0211	77	-0.84	032.42	0
032	Potassium (%)	0123	0.9700	0.0000	1.008	0.0499	0.0211	77	-0.77	032.41	0
032	Potassium (%)	0520	0.9750	0.0100	1.008	0.0499	0.0211	77	-0.67	032.41	0
032	Potassium (%)	0045	0.9755	0.0110	1.008	0.0499	0.0211	77	-0.66	032.42	0
032	Potassium (%)	0870	0.9787	0.0160	1.008	0.0499	0.0211	77	-0.60	032.43	0
032	Potassium (%)	0870	0.9790	0.0124	1.008	0.0499	0.0211	77	-0.59	032.42	0
032	Potassium (%)	0096	0.9846	0.0319	1.008	0.0499	0.0211	77	-0.48	032.42	0
032	Potassium (%)	0144	0.9850	0.0300	1.008	0.0499	0.0211	77	-0.47	032.41	0
032	Potassium (%)	0042	0.9870	0.1060	1.008	0.0499	0.0211	77	-0.43	032.42	0
032	Potassium (%)	0186	0.9916	0.0046	1.008	0.0499	0.0211	77	-0.34	032.52	0
032	Potassium (%)	0868	0.9950	0.0300	1.008	0.0499	0.0211	77	-0.27	032.31	0
032	Potassium (%)	0035	0.9950	0.0300	1.008	0.0499	0.0211	77	-0.27	032.42	0
032	Potassium (%)	0021	0.9950	0.0100	1.008	0.0499	0.0211	77	-0.27	032.43	0
032	Potassium (%)	0164	1.002	0.0160	1.008	0.0499	0.0211	77	-0.13	032.44	0
032	Potassium (%)	0175	1.005	0.0100	1.008	0.0499	0.0211	77	-0.07	032.31	0
032	Potassium (%)	0049	1.005	0.0900	1.008	0.0499	0.0211	77	-0.07	032.41	0
032	Potassium (%)	0229	1.005	0.0100	1.008	0.0499	0.0211	77	-0.07	032.41	0
032	Potassium (%)	0278	1.005	0.0500	1.008	0.0499	0.0211	77	-0.07	032.42	0
032	Potassium (%)	0294	1.005	0.0100	1.008	0.0499	0.0211	77	-0.07	032.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
032	Potassium (%)	0083	1.005	0.0100	1.008	0.0499	0.0211	77	-0.07	032.43	0
032	Potassium (%)	0345	1.005	0.0500	1.008	0.0499	0.0211	77	-0.07	032.43	0
032	Potassium (%)	0590	1.005	0.0300	1.008	0.0499	0.0211	77	-0.07	032.99	0
032	Potassium (%)	0074	1.010	0.0000	1.008	0.0499	0.0211	77	0.03	032.41	0
032	Potassium (%)	0171	1.010	0.0000	1.008	0.0499	0.0211	77	0.03	032.41	0
032	Potassium (%)	0265	1.010	0.1000	1.008	0.0499	0.0211	77	0.03	032.42	0
032	Potassium (%)	0035	1.010	0.0200	1.008	0.0499	0.0211	77	0.03	032.43	0
032	Potassium (%)	0968	1.010	0.0380	1.008	0.0499	0.0211	77	0.03	032.43	0
032	Potassium (%)	0918	1.011	0.0050	1.008	0.0499	0.0211	77	0.04	032.43	0
032	Potassium (%)	0598	1.011	0.0287	1.008	0.0499	0.0211	77	0.05	032.41	0
032	Potassium (%)	0948	1.015	0.0100	1.008	0.0499	0.0211	77	0.13	032.31	0
032	Potassium (%)	0939	1.015	0.0300	1.008	0.0499	0.0211	77	0.13	032.53	0
032	Potassium (%)	0186	1.016	0.0010	1.008	0.0499	0.0211	77	0.14	032.42	0
032	Potassium (%)	0208	1.020	0.0200	1.008	0.0499	0.0211	77	0.23	032.31	0
032	Potassium (%)	0098	1.020	0.0000	1.008	0.0499	0.0211	77	0.23	032.41	0
032	Potassium (%)	0610	1.020	0.0000	1.008	0.0499	0.0211	77	0.23	032.43	0
032	Potassium (%)	0169	1.025	0.0100	1.008	0.0499	0.0211	77	0.33	032.32	0
032	Potassium (%)	0148	1.025	0.0100	1.008	0.0499	0.0211	77	0.33	032.41	0
032	Potassium (%)	0297	1.025	0.0300	1.008	0.0499	0.0211	77	0.33	032.43	0
032	Potassium (%)	0202	1.032	0.0180	1.008	0.0499	0.0211	77	0.47	032.43	0
032	Potassium (%)	0353	1.040	0.0200	1.008	0.0499	0.0211	77	0.63	032.43	0
032	Potassium (%)	0889	1.043	0.0150	1.008	0.0499	0.0211	77	0.68	032.99	0
032	Potassium (%)	0560	1.044	0.0451	1.008	0.0499	0.0211	77	0.71	032.52	0
032	Potassium (%)	0939	1.045	0.0100	1.008	0.0499	0.0211	77	0.73	032.31	0
032	Potassium (%)	0878	1.045	0.0100	1.008	0.0499	0.0211	77	0.73	032.41	0
032	Potassium (%)	0693	1.047	0.0470	1.008	0.0499	0.0211	77	0.76	032.42	0
032	Potassium (%)	0263	1.047	0.0119	1.008	0.0499	0.0211	77	0.77	032.42	0
032	Potassium (%)	0226	1.050	0.0000	1.008	0.0499	0.0211	77	0.83	032.41	0
032	Potassium (%)	0682	1.050	0.0000	1.008	0.0499	0.0211	77	0.83	032.43	0
032	Potassium (%)	2326	1.050	0.0200	1.008	0.0499	0.0211	77	0.83	032.43	0
032	Potassium (%)	0037	1.052	0.0110	1.008	0.0499	0.0211	77	0.86	032.42	0
032	Potassium (%)	0098	1.055	0.0300	1.008	0.0499	0.0211	77	0.93	032.43	0
032	Potassium (%)	0572	1.055	0.0100	1.008	0.0499	0.0211	77	0.93	032.53	0
032	Potassium (%)	0407	1.056	0.0022	1.008	0.0499	0.0211	77	0.96	032.43	0
032	Potassium (%)	2192	1.070	0.0200	1.008	0.0499	0.0211	77	1.24	032.43	0
032	Potassium (%)	0964	1.070	0.0225	1.008	0.0499	0.0211	77	1.24	032.43	0
032	Potassium (%)	0726	1.095	0.0100	1.008	0.0499	0.0211	77	1.74	032.42	0
032	Potassium (%)	0683	1.100	0.0200	1.008	0.0499	0.0211	77	1.84	032.31	0
032	Potassium (%)	0692	1.100	0.0400	1.008	0.0499	0.0211	77	1.84	032.42	0
032	Potassium (%)	0407	1.104	0.0070	1.008	0.0499	0.0211	77	1.92	032.41	0
032	Potassium (%)	0190	1.105	0.0100	1.008	0.0499	0.0211	77	1.94	032.42	0
032	Potassium (%)	0100	1.120	0.0400	1.008	0.0499	0.0211	77	2.24	032.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
032	Potassium (%)	0510	1.130	0.0000	1.008	0.0499	0.0211	77	2.44	032.43	0
032	Potassium (%)	0358	1.095	0.1900	1.008	0.0499	0.0211	77	1.74	032.41	1
032	Potassium (%)	0366	1.195	0.1500	1.008	0.0499	0.0211	77	3.74	032.42	1
033	Salt as chloride (%)	0885	0.3407	0.0251	0.6254	0.0413	0.0136	66	-6.89	033.00	0
033	Salt as chloride (%)	0208	0.3785	0.0050	0.6254	0.0413	0.0136	66	-5.97	033.99	0
033	Salt as chloride (%)	2073	0.4850	0.0100	0.6254	0.0413	0.0136	66	-3.39	033.99	0
033	Salt as chloride (%)	0726	0.5100	0.0000	0.6254	0.0413	0.0136	66	-2.79	033.03	0
033	Salt as chloride (%)	2129	0.5260	0.0900	0.6254	0.0413	0.0136	66	-2.40	033.99	0
033	Salt as chloride (%)	0144	0.5300	0.0000	0.6254	0.0413	0.0136	66	-2.31	033.03	0
033	Salt as chloride (%)	0619	0.5355	0.0010	0.6254	0.0413	0.0136	66	-2.17	033.01	0
033	Salt as chloride (%)	0536	0.5400	0.0400	0.6254	0.0413	0.0136	66	-2.06	033.99	0
033	Salt as chloride (%)	0358	0.5600	0.0400	0.6254	0.0413	0.0136	66	-1.58	033.99	0
033	Salt as chloride (%)	0897	0.5650	0.0300	0.6254	0.0413	0.0136	66	-1.46	033.00	0
033	Salt as chloride (%)	0171	0.5850	0.0100	0.6254	0.0413	0.0136	66	-0.98	033.00	0
033	Salt as chloride (%)	0186	0.5894	0.0001	0.6254	0.0413	0.0136	66	-0.87	033.01	0
033	Salt as chloride (%)	2295	0.5950	0.0100	0.6254	0.0413	0.0136	66	-0.73	033.00	0
033	Salt as chloride (%)	2009	0.5956	0.0106	0.6254	0.0413	0.0136	66	-0.72	033.00	0
033	Salt as chloride (%)	2196	0.6000	0.0000	0.6254	0.0413	0.0136	66	-0.61	033.00	0
033	Salt as chloride (%)	0003	0.6000	0.0000	0.6254	0.0413	0.0136	66	-0.61	033.01	0
033	Salt as chloride (%)	2329	0.6000	0.0000	0.6254	0.0413	0.0136	66	-0.61	033.99	0
033	Salt as chloride (%)	0868	0.6000	0.0800	0.6254	0.0413	0.0136	66	-0.61	033.05	0
033	Salt as chloride (%)	0123	0.6050	0.0100	0.6254	0.0413	0.0136	66	-0.49	033.00	0
033	Salt as chloride (%)	2246	0.6050	0.0300	0.6254	0.0413	0.0136	66	-0.49	033.00	0
033	Salt as chloride (%)	2062	0.6076	0.0053	0.6254	0.0413	0.0136	66	-0.43	033.00	0
033	Salt as chloride (%)	0910	0.6100	0.0000	0.6254	0.0413	0.0136	66	-0.37	033.99	0
033	Salt as chloride (%)	0650	0.6150	0.0100	0.6254	0.0413	0.0136	66	-0.25	033.99	0
033	Salt as chloride (%)	0693	0.6165	0.0450	0.6254	0.0413	0.0136	66	-0.21	033.00	0
033	Salt as chloride (%)	0878	0.6200	0.0400	0.6254	0.0413	0.0136	66	-0.13	033.00	0
033	Salt as chloride (%)	0893	0.6200	0.0200	0.6254	0.0413	0.0136	66	-0.13	033.00	0
033	Salt as chloride (%)	0083	0.6200	0.0000	0.6254	0.0413	0.0136	66	-0.13	033.01	0
033	Salt as chloride (%)	2193	0.6250	0.0100	0.6254	0.0413	0.0136	66	-0.01	033.00	0
033	Salt as chloride (%)	0148	0.6295	0.0030	0.6254	0.0413	0.0136	66	0.10	033.01	0
033	Salt as chloride (%)	0881	0.6300	0.0200	0.6254	0.0413	0.0136	66	0.11	033.00	0
033	Salt as chloride (%)	0004	0.6300	0.0000	0.6254	0.0413	0.0136	66	0.11	033.01	0
033	Salt as chloride (%)	0948	0.6300	0.0000	0.6254	0.0413	0.0136	66	0.11	033.01	0
033	Salt as chloride (%)	0689	0.6300	0.0000	0.6254	0.0413	0.0136	66	0.11	033.05	0
033	Salt as chloride (%)	0045	0.6305	0.0010	0.6254	0.0413	0.0136	66	0.12	033.00	0
033	Salt as chloride (%)	0309	0.6313	0.0127	0.6254	0.0413	0.0136	66	0.14	033.00	0
033	Salt as chloride (%)	0366	0.6350	0.0500	0.6254	0.0413	0.0136	66	0.23	033.00	0
033	Salt as chloride (%)	0001	0.6350	0.0100	0.6254	0.0413	0.0136	66	0.23	033.01	0
033	Salt as chloride (%)	0175	0.6350	0.0100	0.6254	0.0413	0.0136	66	0.23	033.01	0
033	Salt as chloride (%)	0683	0.6350	0.0100	0.6254	0.0413	0.0136	66	0.23	033.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
033	Salt as chloride (%)	0353	0.6400	0.0200	0.6254	0.0413	0.0136	66	0.35	033.00	0
033	Salt as chloride (%)	0884	0.6400	0.0200	0.6254	0.0413	0.0136	66	0.35	033.00	0
033	Salt as chloride (%)	0100	0.6400	0.0000	0.6254	0.0413	0.0136	66	0.35	033.01	0
033	Salt as chloride (%)	0164	0.6400	0.0000	0.6254	0.0413	0.0136	66	0.35	033.01	0
033	Salt as chloride (%)	0226	0.6400	0.0000	0.6254	0.0413	0.0136	66	0.35	033.01	0
033	Salt as chloride (%)	0242	0.6400	0.0000	0.6254	0.0413	0.0136	66	0.35	033.01	0
033	Salt as chloride (%)	0425	0.6400	0.0000	0.6254	0.0413	0.0136	66	0.35	033.01	0
033	Salt as chloride (%)	0407	0.6430	0.0020	0.6254	0.0413	0.0136	66	0.43	033.01	0
033	Salt as chloride (%)	0098	0.6450	0.0100	0.6254	0.0413	0.0136	66	0.47	033.01	0
033	Salt as chloride (%)	2146	0.6450	0.0100	0.6254	0.0413	0.0136	66	0.47	033.01	0
033	Salt as chloride (%)	0019	0.6500	0.0000	0.6254	0.0413	0.0136	66	0.60	033.01	0
033	Salt as chloride (%)	2319	0.6500	0.0000	0.6254	0.0413	0.0136	66	0.60	033.01	0
033	Salt as chloride (%)	0682	0.6500	0.0000	0.6254	0.0413	0.0136	66	0.60	033.99	0
033	Salt as chloride (%)	2109	0.6505	0.0050	0.6254	0.0413	0.0136	66	0.61	033.01	0
033	Salt as chloride (%)	0229	0.6550	0.0100	0.6254	0.0413	0.0136	66	0.72	033.01	0
033	Salt as chloride (%)	0590	0.6550	0.0500	0.6254	0.0413	0.0136	66	0.72	033.01	0
033	Salt as chloride (%)	0964	0.6560	0.0320	0.6254	0.0413	0.0136	66	0.74	033.01	0
033	Salt as chloride (%)	0510	0.6600	0.0000	0.6254	0.0413	0.0136	66	0.84	033.01	0
033	Salt as chloride (%)	0190	0.6600	0.0000	0.6254	0.0413	0.0136	66	0.84	033.03	0
033	Salt as chloride (%)	2181	0.6750	0.0100	0.6254	0.0413	0.0136	66	1.20	033.99	0
033	Salt as chloride (%)	0539	0.6800	0.0200	0.6254	0.0413	0.0136	66	1.32	033.00	0
033	Salt as chloride (%)	0939	0.6900	0.0200	0.6254	0.0413	0.0136	66	1.56	033.01	0
033	Salt as chloride (%)	0511	0.7000	0.0000	0.6254	0.0413	0.0136	66	1.80	033.00	0
033	Salt as chloride (%)	0298	0.7100	0.0200	0.6254	0.0413	0.0136	66	2.05	033.00	0
033	Salt as chloride (%)	0265	0.7800	0.0200	0.6254	0.0413	0.0136	66	3.74	033.03	0
033	Salt as chloride (%)	0278	0.8850	0.0100	0.6254	0.0413	0.0136	66	6.28	033.01	0
033	Salt as chloride (%)	0183	2.200	0.0000	0.6254	0.0413	0.0136	66	38.08	033.00	0
033	Salt as chloride (%)	0505	0.4500	0.1000	0.6254	0.0413	0.0136	66	-4.24	033.03	1
033	Salt as chloride (%)	2076	0.7000	0.1400	0.6254	0.0413	0.0136	66	1.80	033.00	1
033	Salt as chloride (%)	2043	9.660	0.0400	0.6254	0.0413	0.0136	66	218.50	033.05	2
034	Selenium (ppm)	0045	0.3435	0.0030	0.4420	0.0804	0.0425	15	-1.23	034.04	0
034	Selenium (ppm)	0164	0.3470	0.0580	0.4420	0.0804	0.0425	15	-1.18	034.53	0
034	Selenium (ppm)	0171	0.3705	0.0010	0.4420	0.0804	0.0425	15	-0.89	034.04	0
034	Selenium (ppm)	0407	0.3800	0.0200	0.4420	0.0804	0.0425	15	-0.77	034.41	0
034	Selenium (ppm)	0021	0.3900	0.0400	0.4420	0.0804	0.0425	15	-0.65	034.53	0
034	Selenium (ppm)	0096	0.3925	0.0575	0.4420	0.0804	0.0425	15	-0.62	034.52	0
034	Selenium (ppm)	0169	0.4450	0.0300	0.4420	0.0804	0.0425	15	0.04	034.04	0
034	Selenium (ppm)	0186	0.4450	0.0060	0.4420	0.0804	0.0425	15	0.04	034.52	0
034	Selenium (ppm)	0553	0.4455	0.0210	0.4420	0.0804	0.0425	15	0.04	034.53	0
034	Selenium (ppm)	0910	0.4750	0.1900	0.4420	0.0804	0.0425	15	0.41	034.52	0
034	Selenium (ppm)	0918	0.4795	0.0210	0.4420	0.0804	0.0425	15	0.47	034.53	0
034	Selenium (ppm)	0560	0.4860	0.0400	0.4420	0.0804	0.0425	15	0.55	034.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
034	Selenium (ppm)	0939	0.5050	0.0700	0.4420	0.0804	0.0425	15	0.78	034.53	0
034	Selenium (ppm)	0968	0.8290	0.0400	0.4420	0.0804	0.0425	15	4.82	034.43	0
034	Selenium (ppm)	0563	0.9361	0.0400	0.4420	0.0804	0.0425	15	6.15	034.04	0
034	Selenium (ppm)	0407	0.9550	0.7500	0.4420	0.0804	0.0425	15	6.38	034.43	1
034	Selenium (ppm)	0692	0.0000	0.0000	0.4420	0.0804	0.0425	15		034.42	4
034	Selenium (ppm)	2302	< 0.01		0.4420	0.0804	0.0425	15		034.99	5
034	Selenium (ppm)	0682	< 5		0.4420	0.0804	0.0425	15		034.43	5
034	Selenium (ppm)	0096	< 20		0.4420	0.0804	0.0425	15		034.42	5
034	Selenium (ppm)	0021	< 22		0.4420	0.0804	0.0425	15		034.43	5
035	Sodium (%)	0889	0.0700	0.0000	0.2076	0.0134	0.0072	93	-10.28	035.99	0
035	Sodium (%)	0536	0.1650	0.0100	0.2076	0.0134	0.0072	93	-3.18	035.31	0
035	Sodium (%)	0939	0.1750	0.0300	0.2076	0.0134	0.0072	93	-2.44	035.53	0
035	Sodium (%)	0154	0.1772	0.0034	0.2076	0.0134	0.0072	93	-2.27	035.52	0
035	Sodium (%)	0175	0.1850	0.0100	0.2076	0.0134	0.0072	93	-1.69	035.31	0
035	Sodium (%)	0910	0.1850	0.0100	0.2076	0.0134	0.0072	93	-1.69	035.41	0
035	Sodium (%)	2129	0.1865	0.0010	0.2076	0.0134	0.0072	93	-1.58	035.42	0
035	Sodium (%)	0096	0.1890	0.0017	0.2076	0.0134	0.0072	93	-1.39	035.42	0
035	Sodium (%)	2246	0.1900	0.0000	0.2076	0.0134	0.0072	93	-1.32	035.31	0
035	Sodium (%)	0278	0.1900	0.0200	0.2076	0.0134	0.0072	93	-1.32	035.42	0
035	Sodium (%)	0510	0.1930	0.0020	0.2076	0.0134	0.0072	93	-1.09	035.43	0
035	Sodium (%)	0563	0.1949	0.0013	0.2076	0.0134	0.0072	93	-0.95	035.31	0
035	Sodium (%)	0169	0.1950	0.0300	0.2076	0.0134	0.0072	93	-0.94	035.32	0
035	Sodium (%)	2143	0.1950	0.0100	0.2076	0.0134	0.0072	93	-0.94	035.43	0
035	Sodium (%)	0208	0.1965	0.0130	0.2076	0.0134	0.0072	93	-0.83	035.41	0
035	Sodium (%)	0553	0.1965	0.0050	0.2076	0.0134	0.0072	93	-0.83	035.53	0
035	Sodium (%)	0357	0.1966	0.0064	0.2076	0.0134	0.0072	93	-0.82	035.42	0
035	Sodium (%)	0042	0.1990	0.0260	0.2076	0.0134	0.0072	93	-0.64	035.42	0
035	Sodium (%)	0045	0.1990	0.0080	0.2076	0.0134	0.0072	93	-0.64	035.42	0
035	Sodium (%)	0560	0.1992	0.0062	0.2076	0.0134	0.0072	93	-0.63	035.52	0
035	Sodium (%)	0870	0.1999	0.0029	0.2076	0.0134	0.0072	93	-0.58	035.43	0
035	Sodium (%)	0683	0.2000	0.0000	0.2076	0.0134	0.0072	93	-0.57	035.31	0
035	Sodium (%)	0019	0.2000	0.0000	0.2076	0.0134	0.0072	93	-0.57	035.41	0
035	Sodium (%)	0229	0.2000	0.0000	0.2076	0.0134	0.0072	93	-0.57	035.41	0
035	Sodium (%)	0298	0.2000	0.0200	0.2076	0.0134	0.0072	93	-0.57	035.41	0
035	Sodium (%)	0265	0.2000	0.0200	0.2076	0.0134	0.0072	93	-0.57	035.42	0
035	Sodium (%)	0294	0.2000	0.0000	0.2076	0.0134	0.0072	93	-0.57	035.42	0
035	Sodium (%)	0021	0.2000	0.0200	0.2076	0.0134	0.0072	93	-0.57	035.43	0
035	Sodium (%)	2234	0.2000	0.0000	0.2076	0.0134	0.0072	93	-0.57	035.43	0
035	Sodium (%)	0242	0.2000	0.0000	0.2076	0.0134	0.0072	93	-0.57	035.99	0
035	Sodium (%)	0033	0.2005	0.0010	0.2076	0.0134	0.0072	93	-0.53	035.43	0
035	Sodium (%)	0096	0.2008	0.0162	0.2076	0.0134	0.0072	93	-0.51	035.52	0
035	Sodium (%)	2319	0.2015	0.0050	0.2076	0.0134	0.0072	93	-0.46	035.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
035	Sodium (%)	0918	0.2015	0.0010	0.2076	0.0134	0.0072	93	-0.46	035.43	0
035	Sodium (%)	0263	0.2020	0.0020	0.2076	0.0134	0.0072	93	-0.42	035.42	0
035	Sodium (%)	0870	0.2023	0.0062	0.2076	0.0134	0.0072	93	-0.40	035.42	0
035	Sodium (%)	0366	0.2030	0.0280	0.2076	0.0134	0.0072	93	-0.34	035.42	0
035	Sodium (%)	0164	0.2035	0.0010	0.2076	0.0134	0.0072	93	-0.31	035.41	0
035	Sodium (%)	0529	0.2040	0.0000	0.2076	0.0134	0.0072	93	-0.27	035.31	0
035	Sodium (%)	0300	0.2041	0.0046	0.2076	0.0134	0.0072	93	-0.26	035.43	0
035	Sodium (%)	0353	0.2043	0.0150	0.2076	0.0134	0.0072	93	-0.25	035.43	0
035	Sodium (%)	0345	0.2045	0.0030	0.2076	0.0134	0.0072	93	-0.23	035.43	0
035	Sodium (%)	0939	0.2050	0.0100	0.2076	0.0134	0.0072	93	-0.20	035.31	0
035	Sodium (%)	0049	0.2050	0.0100	0.2076	0.0134	0.0072	93	-0.20	035.41	0
035	Sodium (%)	0144	0.2050	0.0100	0.2076	0.0134	0.0072	93	-0.20	035.41	0
035	Sodium (%)	0035	0.2050	0.0100	0.2076	0.0134	0.0072	93	-0.20	035.42	0
035	Sodium (%)	0297	0.2050	0.0100	0.2076	0.0134	0.0072	93	-0.20	035.43	0
035	Sodium (%)	0505	0.2050	0.0300	0.2076	0.0134	0.0072	93	-0.20	035.43	0
035	Sodium (%)	0520	0.2064	0.0021	0.2076	0.0134	0.0072	93	-0.09	035.41	0
035	Sodium (%)	0148	0.2065	0.0030	0.2076	0.0134	0.0072	93	-0.08	035.41	0
035	Sodium (%)	0948	0.2100	0.0000	0.2076	0.0134	0.0072	93	0.18	035.31	0
035	Sodium (%)	0098	0.2100	0.0000	0.2076	0.0134	0.0072	93	0.18	035.41	0
035	Sodium (%)	0171	0.2100	0.0000	0.2076	0.0134	0.0072	93	0.18	035.41	0
035	Sodium (%)	0878	0.2100	0.0000	0.2076	0.0134	0.0072	93	0.18	035.41	0
035	Sodium (%)	0190	0.2100	0.0000	0.2076	0.0134	0.0072	93	0.18	035.42	0
035	Sodium (%)	0726	0.2100	0.0000	0.2076	0.0134	0.0072	93	0.18	035.42	0
035	Sodium (%)	0035	0.2100	0.0000	0.2076	0.0134	0.0072	93	0.18	035.43	0
035	Sodium (%)	0083	0.2100	0.0000	0.2076	0.0134	0.0072	93	0.18	035.43	0
035	Sodium (%)	0425	0.2100	0.0200	0.2076	0.0134	0.0072	93	0.18	035.43	0
035	Sodium (%)	0693	0.2115	0.0050	0.2076	0.0134	0.0072	93	0.29	035.42	0
035	Sodium (%)	0037	0.2125	0.0050	0.2076	0.0134	0.0072	93	0.37	035.42	0
035	Sodium (%)	0142	0.2131	0.0028	0.2076	0.0134	0.0072	93	0.41	035.31	0
035	Sodium (%)	0868	0.2135	0.0010	0.2076	0.0134	0.0072	93	0.44	035.01	0
035	Sodium (%)	0572	0.2145	0.0030	0.2076	0.0134	0.0072	93	0.51	035.53	0
035	Sodium (%)	0123	0.2150	0.0100	0.2076	0.0134	0.0072	93	0.55	035.41	0
035	Sodium (%)	0226	0.2150	0.0100	0.2076	0.0134	0.0072	93	0.55	035.41	0
035	Sodium (%)	0098	0.2150	0.0100	0.2076	0.0134	0.0072	93	0.55	035.43	0
035	Sodium (%)	0610	0.2150	0.0100	0.2076	0.0134	0.0072	93	0.55	035.43	0
035	Sodium (%)	2089	0.2150	0.0100	0.2076	0.0134	0.0072	93	0.55	035.43	0
035	Sodium (%)	0590	0.2150	0.0100	0.2076	0.0134	0.0072	93	0.55	035.99	0
035	Sodium (%)	2146	0.2160	0.0080	0.2076	0.0134	0.0072	93	0.63	035.31	0
035	Sodium (%)	0202	0.2160	0.0000	0.2076	0.0134	0.0072	93	0.63	035.43	0
035	Sodium (%)	2326	0.2160	0.0040	0.2076	0.0134	0.0072	93	0.63	035.43	0
035	Sodium (%)	0964	0.2172	0.0024	0.2076	0.0134	0.0072	93	0.72	035.43	0
035	Sodium (%)	0407	0.2200	0.0041	0.2076	0.0134	0.0072	93	0.92	035.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
035	Sodium (%)	0358	0.2200	0.0200	0.2076	0.0134	0.0072	93	0.93	035.41	0
035	Sodium (%)	0010	0.2200	0.0000	0.2076	0.0134	0.0072	93	0.93	035.43	0
035	Sodium (%)	0682	0.2200	0.0000	0.2076	0.0134	0.0072	93	0.93	035.43	0
035	Sodium (%)	0186	0.2216	0.0018	0.2076	0.0134	0.0072	93	1.04	035.52	0
035	Sodium (%)	0066	0.2240	0.0140	0.2076	0.0134	0.0072	93	1.22	035.41	0
035	Sodium (%)	0868	0.2250	0.0100	0.2076	0.0134	0.0072	93	1.30	035.31	0
035	Sodium (%)	0100	0.2250	0.0100	0.2076	0.0134	0.0072	93	1.30	035.99	0
035	Sodium (%)	0598	0.2256	0.0016	0.2076	0.0134	0.0072	93	1.34	035.41	0
035	Sodium (%)	0407	0.2262	0.0083	0.2076	0.0134	0.0072	93	1.38	035.41	0
035	Sodium (%)	0619	0.2295	0.0090	0.2076	0.0134	0.0072	93	1.63	035.41	0
035	Sodium (%)	2196	0.2310	0.0000	0.2076	0.0134	0.0072	93	1.75	035.31	0
035	Sodium (%)	0004	0.2350	0.0300	0.2076	0.0134	0.0072	93	2.05	035.41	0
035	Sodium (%)	0692	0.2400	0.0000	0.2076	0.0134	0.0072	93	2.42	035.42	0
035	Sodium (%)	2302	0.2400	0.0000	0.2076	0.0134	0.0072	93	2.42	035.99	0
035	Sodium (%)	0186	0.2463	0.0022	0.2076	0.0134	0.0072	93	2.89	035.42	0
035	Sodium (%)	0089	0.2600	0.0000	0.2076	0.0134	0.0072	93	3.91	035.43	0
035	Sodium (%)	0689	0.2800	0.0200	0.2076	0.0134	0.0072	93	5.41	035.41	0
035	Sodium (%)	0656	0.3500	0.0000	0.2076	0.0134	0.0072	93	10.63	035.31	0
035	Sodium (%)	0108	0.1800	0.0400	0.2076	0.0134	0.0072	93	-2.06	035.05	1
036	Sulfur (%)	2129	0.2035	0.0070	0.2606	0.0214	0.0077	49	-2.67	036.42	0
036	Sulfur (%)	0229	0.2250	0.0100	0.2606	0.0214	0.0077	49	-1.67	036.42	0
036	Sulfur (%)	0889	0.2275	0.0050	0.2606	0.0214	0.0077	49	-1.55	036.99	0
036	Sulfur (%)	0242	0.2300	0.0000	0.2606	0.0214	0.0077	49	-1.43	036.99	0
036	Sulfur (%)	0870	0.2319	0.0005	0.2606	0.0214	0.0077	49	-1.35	036.42	0
036	Sulfur (%)	0407	0.2340	0.0071	0.2606	0.0214	0.0077	49	-1.25	036.43	0
036	Sulfur (%)	0229	0.2350	0.0100	0.2606	0.0214	0.0077	49	-1.20	036.04	0
036	Sulfur (%)	0045	0.2385	0.0050	0.2606	0.0214	0.0077	49	-1.03	036.42	0
036	Sulfur (%)	0278	0.2400	0.0200	0.2606	0.0214	0.0077	49	-0.96	036.42	0
036	Sulfur (%)	0870	0.2447	0.0012	0.2606	0.0214	0.0077	49	-0.74	036.43	0
036	Sulfur (%)	0353	0.2450	0.0100	0.2606	0.0214	0.0077	49	-0.73	036.43	0
036	Sulfur (%)	0096	0.2453	0.0085	0.2606	0.0214	0.0077	49	-0.72	036.42	0
036	Sulfur (%)	0910	0.2471	0.0209	0.2606	0.0214	0.0077	49	-0.63	036.42	0
036	Sulfur (%)	0910	0.2471	0.0209	0.2606	0.0214	0.0077	49	-0.63	036.52	0
036	Sulfur (%)	0265	0.2500	0.0000	0.2606	0.0214	0.0077	49	-0.50	036.42	0
036	Sulfur (%)	0083	0.2500	0.0000	0.2606	0.0214	0.0077	49	-0.50	036.43	0
036	Sulfur (%)	0297	0.2500	0.0000	0.2606	0.0214	0.0077	49	-0.50	036.43	0
036	Sulfur (%)	2326	0.2500	0.0020	0.2606	0.0214	0.0077	49	-0.50	036.43	0
036	Sulfur (%)	0357	0.2540	0.0003	0.2606	0.0214	0.0077	49	-0.31	036.42	0
036	Sulfur (%)	0294	0.2550	0.0100	0.2606	0.0214	0.0077	49	-0.26	036.42	0
036	Sulfur (%)	0692	0.2550	0.0300	0.2606	0.0214	0.0077	49	-0.26	036.42	0
036	Sulfur (%)	0693	0.2575	0.0090	0.2606	0.0214	0.0077	49	-0.14	036.42	0
036	Sulfur (%)	0407	0.2597	0.0012	0.2606	0.0214	0.0077	49	-0.04	036.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
036	Sulfur (%)	0226	0.2600	0.0000	0.2606	0.0214	0.0077	49	-0.03	036.04	0
036	Sulfur (%)	0042	0.2620	0.0280	0.2606	0.0214	0.0077	49	0.07	036.42	0
036	Sulfur (%)	0345	0.2625	0.0050	0.2606	0.0214	0.0077	49	0.09	036.43	0
036	Sulfur (%)	0164	0.2650	0.0100	0.2606	0.0214	0.0077	49	0.21	036.42	0
036	Sulfur (%)	0035	0.2650	0.0100	0.2606	0.0214	0.0077	49	0.21	036.43	0
036	Sulfur (%)	0553	0.2680	0.0280	0.2606	0.0214	0.0077	49	0.35	036.53	0
036	Sulfur (%)	0186	0.2688	0.0011	0.2606	0.0214	0.0077	49	0.38	036.52	0
036	Sulfur (%)	0918	0.2695	0.0010	0.2606	0.0214	0.0077	49	0.42	036.43	0
036	Sulfur (%)	0035	0.2700	0.0000	0.2606	0.0214	0.0077	49	0.44	036.42	0
036	Sulfur (%)	0171	0.2700	0.0000	0.2606	0.0214	0.0077	49	0.44	036.42	0
036	Sulfur (%)	0708	0.2700	0.0040	0.2606	0.0214	0.0077	49	0.44	036.42	0
036	Sulfur (%)	0726	0.2700	0.0000	0.2606	0.0214	0.0077	49	0.44	036.42	0
036	Sulfur (%)	0148	0.2705	0.0110	0.2606	0.0214	0.0077	49	0.46	036.04	0
036	Sulfur (%)	0190	0.2750	0.0100	0.2606	0.0214	0.0077	49	0.67	036.42	0
036	Sulfur (%)	0169	0.2750	0.0100	0.2606	0.0214	0.0077	49	0.67	036.43	0
036	Sulfur (%)	0186	0.2796	0.0084	0.2606	0.0214	0.0077	49	0.89	036.42	0
036	Sulfur (%)	0021	0.2800	0.0000	0.2606	0.0214	0.0077	49	0.91	036.43	0
036	Sulfur (%)	0098	0.2811	0.0062	0.2606	0.0214	0.0077	49	0.96	036.43	0
036	Sulfur (%)	0202	0.2820	0.0100	0.2606	0.0214	0.0077	49	1.00	036.43	0
036	Sulfur (%)	0505	0.2850	0.0100	0.2606	0.0214	0.0077	49	1.14	036.43	0
036	Sulfur (%)	0964	0.2856	0.0053	0.2606	0.0214	0.0077	49	1.17	036.43	0
036	Sulfur (%)	0560	0.2863	0.0309	0.2606	0.0214	0.0077	49	1.20	036.52	0
036	Sulfur (%)	0037	0.2875	0.0010	0.2606	0.0214	0.0077	49	1.26	036.42	0
036	Sulfur (%)	0510	0.2900	0.0000	0.2606	0.0214	0.0077	49	1.38	036.43	0
036	Sulfur (%)	0682	0.3000	0.0000	0.2606	0.0214	0.0077	49	1.85	036.43	0
036	Sulfur (%)	2089	0.3250	0.0100	0.2606	0.0214	0.0077	49	3.02	036.43	0
036	Sulfur (%)	0366	0.2700	0.0400	0.2606	0.0214	0.0077	49	0.44	036.42	1
037	Zinc (ppm)	2234	16.60	1.220	105.1	7.932	3.773	89	-11.16	037.43	0
037	Zinc (ppm)	2246	79.36	1.280	105.1	7.932	3.773	89	-3.25	037.31	0
037	Zinc (ppm)	0047	81.30	2.200	105.1	7.932	3.773	89	-3.00	037.52	0
037	Zinc (ppm)	0066	87.74	0.9390	105.1	7.932	3.773	89	-2.19	037.44	0
037	Zinc (ppm)	0536	88.03	3.350	105.1	7.932	3.773	89	-2.15	037.31	0
037	Zinc (ppm)	0009	89.50	1.000	105.1	7.932	3.773	89	-1.97	037.42	0
037	Zinc (ppm)	2129	90.10	1.740	105.1	7.932	3.773	89	-1.89	037.42	0
037	Zinc (ppm)	0910	91.50	3.000	105.1	7.932	3.773	89	-1.72	037.41	0
037	Zinc (ppm)	0045	93.50	6.600	105.1	7.932	3.773	89	-1.46	037.42	0
037	Zinc (ppm)	0242	94.50	3.000	105.1	7.932	3.773	89	-1.34	037.99	0
037	Zinc (ppm)	0263	94.52	0.9500	105.1	7.932	3.773	89	-1.34	037.42	0
037	Zinc (ppm)	0918	95.01	2.860	105.1	7.932	3.773	89	-1.27	037.43	0
037	Zinc (ppm)	0889	97.38	2.715	105.1	7.932	3.773	89	-0.97	037.99	0
037	Zinc (ppm)	2143	97.44	4.120	105.1	7.932	3.773	89	-0.97	037.43	0
037	Zinc (ppm)	0297	97.50	5.000	105.1	7.932	3.773	89	-0.96	037.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
037	Zinc (ppm)	0425	97.70	5.400	105.1	7.932	3.773	89	-0.93	037.43	0
037	Zinc (ppm)	0186	98.00	8.000	105.1	7.932	3.773	89	-0.90	037.42	0
037	Zinc (ppm)	0939	98.83	4.340	105.1	7.932	3.773	89	-0.79	037.53	0
037	Zinc (ppm)	0208	99.01	7.990	105.1	7.932	3.773	89	-0.77	037.41	0
037	Zinc (ppm)	0610	99.05	4.100	105.1	7.932	3.773	89	-0.76	037.43	0
037	Zinc (ppm)	0590	99.70	0.4000	105.1	7.932	3.773	89	-0.68	037.99	0
037	Zinc (ppm)	0083	100.5	1.000	105.1	7.932	3.773	89	-0.58	037.43	0
037	Zinc (ppm)	0357	100.7	0.0840	105.1	7.932	3.773	89	-0.56	037.42	0
037	Zinc (ppm)	0226	100.8	0.1300	105.1	7.932	3.773	89	-0.54	037.41	0
037	Zinc (ppm)	0294	101.5	1.100	105.1	7.932	3.773	89	-0.46	037.42	0
037	Zinc (ppm)	0968	101.5	1.000	105.1	7.932	3.773	89	-0.46	037.43	0
037	Zinc (ppm)	0529	101.6	0.1000	105.1	7.932	3.773	89	-0.45	037.31	0
037	Zinc (ppm)	2196	101.6	0.0000	105.1	7.932	3.773	89	-0.44	037.31	0
037	Zinc (ppm)	0345	101.6	2.750	105.1	7.932	3.773	89	-0.44	037.43	0
037	Zinc (ppm)	0265	101.7	4.700	105.1	7.932	3.773	89	-0.44	037.42	0
037	Zinc (ppm)	0049	102.4	3.590	105.1	7.932	3.773	89	-0.34	037.41	0
037	Zinc (ppm)	0870	102.7	0.7000	105.1	7.932	3.773	89	-0.31	037.42	0
037	Zinc (ppm)	0096	102.9	9.118	105.1	7.932	3.773	89	-0.28	037.52	0
037	Zinc (ppm)	0563	103.0	1.755	105.1	7.932	3.773	89	-0.26	037.31	0
037	Zinc (ppm)	0035	103.1	3.400	105.1	7.932	3.773	89	-0.25	037.43	0
037	Zinc (ppm)	0870	103.2	1.100	105.1	7.932	3.773	89	-0.25	037.43	0
037	Zinc (ppm)	0689	103.7	0.7000	105.1	7.932	3.773	89	-0.18	037.31	0
037	Zinc (ppm)	0560	104.1	1.500	105.1	7.932	3.773	89	-0.13	037.52	0
037	Zinc (ppm)	0229	104.6	5.100	105.1	7.932	3.773	89	-0.07	037.41	0
037	Zinc (ppm)	0021	105.0	10.00	105.1	7.932	3.773	89	-0.01	037.43	0
037	Zinc (ppm)	0169	105.0	2.000	105.1	7.932	3.773	89	-0.01	037.43	0
037	Zinc (ppm)	0164	105.0	2.000	105.1	7.932	3.773	89	-0.01	037.44	0
037	Zinc (ppm)	0171	105.1	0.1000	105.1	7.932	3.773	89	-0.01	037.41	0
037	Zinc (ppm)	0939	105.3	1.820	105.1	7.932	3.773	89	0.02	037.31	0
037	Zinc (ppm)	0619	105.5	9.000	105.1	7.932	3.773	89	0.05	037.41	0
037	Zinc (ppm)	0278	105.6	2.700	105.1	7.932	3.773	89	0.06	037.42	0
037	Zinc (ppm)	0096	105.6	13.59	105.1	7.932	3.773	89	0.06	037.42	0
037	Zinc (ppm)	0098	105.7	2.300	105.1	7.932	3.773	89	0.07	037.41	0
037	Zinc (ppm)	0656	106.0	0.0000	105.1	7.932	3.773	89	0.11	037.31	0
037	Zinc (ppm)	0019	106.3	1.400	105.1	7.932	3.773	89	0.15	037.41	0
037	Zinc (ppm)	0358	106.5	5.310	105.1	7.932	3.773	89	0.17	037.41	0
037	Zinc (ppm)	0520	106.5	7.000	105.1	7.932	3.773	89	0.18	037.41	0
037	Zinc (ppm)	0042	106.5	9.000	105.1	7.932	3.773	89	0.18	037.42	0
037	Zinc (ppm)	0510	106.5	1.000	105.1	7.932	3.773	89	0.18	037.43	0
037	Zinc (ppm)	0505	107.2	0.4000	105.1	7.932	3.773	89	0.26	037.43	0
037	Zinc (ppm)	0693	107.5	1.572	105.1	7.932	3.773	89	0.30	037.42	0
037	Zinc (ppm)	0366	107.5	17.00	105.1	7.932	3.773	89	0.30	037.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
037	Zinc (ppm)	2326	107.5	5.000	105.1	7.932	3.773	89	0.30	037.43	0
037	Zinc (ppm)	0553	107.5	1.000	105.1	7.932	3.773	89	0.30	037.53	0
037	Zinc (ppm)	0598	107.8	0.7000	105.1	7.932	3.773	89	0.33	037.41	0
037	Zinc (ppm)	2188	108.3	0.7900	105.1	7.932	3.773	89	0.40	037.31	0
037	Zinc (ppm)	0948	108.5	1.200	105.1	7.932	3.773	89	0.42	037.33	0
037	Zinc (ppm)	0868	108.5	11.00	105.1	7.932	3.773	89	0.43	037.31	0
037	Zinc (ppm)	0878	108.5	3.000	105.1	7.932	3.773	89	0.43	037.41	0
037	Zinc (ppm)	0186	108.5	5.000	105.1	7.932	3.773	89	0.43	037.52	0
037	Zinc (ppm)	0964	108.9	8.533	105.1	7.932	3.773	89	0.48	037.43	0
037	Zinc (ppm)	2192	109.5	1.000	105.1	7.932	3.773	89	0.55	037.43	0
037	Zinc (ppm)	0202	110.2	7.500	105.1	7.932	3.773	89	0.64	037.43	0
037	Zinc (ppm)	0175	111.0	2.000	105.1	7.932	3.773	89	0.74	037.31	0
037	Zinc (ppm)	0726	111.1	0.1000	105.1	7.932	3.773	89	0.75	037.42	0
037	Zinc (ppm)	0682	111.1	0.0000	105.1	7.932	3.773	89	0.75	037.43	0
037	Zinc (ppm)	0572	112.5	1.000	105.1	7.932	3.773	89	0.93	037.53	0
037	Zinc (ppm)	0407	112.6	0.2107	105.1	7.932	3.773	89	0.94	037.41	0
037	Zinc (ppm)	0027	112.9	9.992	105.1	7.932	3.773	89	0.98	037.43	0
037	Zinc (ppm)	0074	113.0	12.00	105.1	7.932	3.773	89	0.99	037.41	0
037	Zinc (ppm)	0098	113.2	3.900	105.1	7.932	3.773	89	1.01	037.43	0
037	Zinc (ppm)	0035	113.4	1.700	105.1	7.932	3.773	89	1.04	037.42	0
037	Zinc (ppm)	0692	113.5	17.00	105.1	7.932	3.773	89	1.06	037.42	0
037	Zinc (ppm)	0407	113.9	5.327	105.1	7.932	3.773	89	1.11	037.43	0
037	Zinc (ppm)	0208	114.0	0.0000	105.1	7.932	3.773	89	1.12	037.31	0
037	Zinc (ppm)	0148	116.4	9.100	105.1	7.932	3.773	89	1.42	037.41	0
037	Zinc (ppm)	0190	117.1	0.4000	105.1	7.932	3.773	89	1.51	037.42	0
037	Zinc (ppm)	0004	117.6	4.170	105.1	7.932	3.773	89	1.58	037.41	0
037	Zinc (ppm)	0037	119.3	0.8000	105.1	7.932	3.773	89	1.79	037.42	0
037	Zinc (ppm)	2089	119.3	1.170	105.1	7.932	3.773	89	1.79	037.43	0
037	Zinc (ppm)	0033	120.0	16.00	105.1	7.932	3.773	89	1.88	037.43	0
037	Zinc (ppm)	0353	123.5	9.000	105.1	7.932	3.773	89	2.32	037.43	0
037	Zinc (ppm)	2302	125.0	0.0000	105.1	7.932	3.773	89	2.51	037.99	0
037	Zinc (ppm)	0100	127.5	3.000	105.1	7.932	3.773	89	2.82	037.99	0
037	Zinc (ppm)	2319	10.06	0.4800	105.1	7.932	3.773	89	-11.98	037.42	2
038	Molybdenum (ppm)	0171	1.086	0.0070	1.511	0.1964	0.1231	22	-2.16	038.41	0
038	Molybdenum (ppm)	0353	1.139	0.0132	1.511	0.1964	0.1231	22	-1.89	038.43	0
038	Molybdenum (ppm)	0939	1.290	0.1000	1.511	0.1964	0.1231	22	-1.12	038.53	0
038	Molybdenum (ppm)	0345	1.350	0.0800	1.511	0.1964	0.1231	22	-0.82	038.43	0
038	Molybdenum (ppm)	0407	1.411	0.0050	1.511	0.1964	0.1231	22	-0.51	038.41	0
038	Molybdenum (ppm)	0918	1.417	0.0080	1.511	0.1964	0.1231	22	-0.48	038.53	0
038	Molybdenum (ppm)	0910	1.420	0.0600	1.511	0.1964	0.1231	22	-0.46	038.52	0
038	Molybdenum (ppm)	0693	1.442	0.1780	1.511	0.1964	0.1231	22	-0.35	038.42	0
038	Molybdenum (ppm)	0510	1.450	0.1000	1.511	0.1964	0.1231	22	-0.31	038.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
038	Molybdenum (ppm)	0407	1.454	0.0291	1.511	0.1964	0.1231	22	-0.29	038.43	0
038	Molybdenum (ppm)	0164	1.455	0.0100	1.511	0.1964	0.1231	22	-0.28	038.53	0
038	Molybdenum (ppm)	0964	1.518	0.0350	1.511	0.1964	0.1231	22	0.04	038.43	0
038	Molybdenum (ppm)	0096	1.534	0.0530	1.511	0.1964	0.1231	22	0.12	038.52	0
038	Molybdenum (ppm)	0226	1.535	0.0300	1.511	0.1964	0.1231	22	0.12	038.41	0
038	Molybdenum (ppm)	0278	1.535	0.1300	1.511	0.1964	0.1231	22	0.12	038.42	0
038	Molybdenum (ppm)	0083	1.600	0.0000	1.511	0.1964	0.1231	22	0.46	038.43	0
038	Molybdenum (ppm)	0553	1.605	0.0900	1.511	0.1964	0.1231	22	0.48	038.53	0
038	Molybdenum (ppm)	0169	1.670	0.4000	1.511	0.1964	0.1231	22	0.81	038.43	0
038	Molybdenum (ppm)	0021	1.700	0.2000	1.511	0.1964	0.1231	22	0.96	038.53	0
038	Molybdenum (ppm)	0045	2.065	0.5100	1.511	0.1964	0.1231	22	2.82	038.42	0
038	Molybdenum (ppm)	0265	2.460	0.4800	1.511	0.1964	0.1231	22	4.83	038.42	0
038	Molybdenum (ppm)	0297	2.475	0.1900	1.511	0.1964	0.1231	22	4.91	038.43	0
038	Molybdenum (ppm)	0096	< 2		1.511	0.1964	0.1231	22		038.42	5
038	Molybdenum (ppm)	0021	< 4		1.511	0.1964	0.1231	22		038.43	5
038	Molybdenum (ppm)	0186	< 5		1.511	0.1964	0.1231	22		038.52	5
040	Barium (ppm)	0918	5.734	0.0140				3		040.53	0
040	Barium (ppm)	0096	5.929	0.2523				3		040.52	0
040	Barium (ppm)	0560	7.055	0.1900				3		040.52	0
041	Vanadium (ppm)	0553	0.5815	0.0290				2		041.53	0
041	Vanadium (ppm)	0021	0.7200	0.1000				2		041.53	0
041	Vanadium (ppm)	0021	< 2					2		041.43	5
042	Chloride (%)	0948	0.3800	0.0000	0.5890	0.2687	0.0140	5		042.00	0
042	Chloride (%)	2326	0.3900	0.0000	0.5890	0.2687	0.0140	5		042.00	0
042	Chloride (%)	0297	0.4200	0.0400	0.5890	0.2687	0.0140	5		042.99	0
042	Chloride (%)	2199	0.8050	0.0100	0.5890	0.2687	0.0140	5		042.00	0
042	Chloride (%)	2200	0.9500	0.0200	0.5890	0.2687	0.0140	5		042.00	0
101	Choline Chloride (ppm)	0227	2,050	20.00				1		101.99	0
102	Niacin (ppm)	0160	25.10	10.00	66.96	32.39	5.925	4		102.99	0
102	Niacin (ppm)	0010	67.00	7.000	66.96	32.39	5.925	4		102.02	0
102	Niacin (ppm)	2319	71.75	0.7000	66.96	32.39	5.925	4		102.02	0
102	Niacin (ppm)	0227	104.0	6.000	66.96	32.39	5.925	4		102.01	0
103	Pantothenic Acid (ppm)	0227	19.45	0.1000				3		103.01	0
103	Pantothenic Acid (ppm)	0010	23.45	0.1000				3		103.02	0
103	Pantothenic Acid (ppm)	0160	46.25	32.30				3		103.99	0
104	Riboflavin (ppm)	0910	13.17	0.2500				3		104.03	0
104	Riboflavin (ppm)	0227	16.25	1.500				3		104.00	0
104	Riboflavin (ppm)	0171	21.20	0.0000				3		104.00	0
105	Thiamine (ppm)	0010	2.550	0.5000				3		105.00	0
105	Thiamine (ppm)	0910	3.080	0.0600				3		105.00	0
105	Thiamine (ppm)	0227	5.135	0.0300				3		105.01	0
106	Vitamin A (KU / kg)	0227	1.320	0.0000	13.52	2.736	1.137	11	-4.46	106.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
106	Vitamin A (KU / kg)	0921	5.096	0.0000	13.52	2.736	1.137	11	-3.08	106.02	0
106	Vitamin A (KU / kg)	0905	11.76	2.310	13.52	2.736	1.137	11	-0.64	106.02	0
106	Vitamin A (KU / kg)	0619	13.50	3.000	13.52	2.736	1.137	11	-0.01	106.02	0
106	Vitamin A (KU / kg)	2319	13.55	0.1000	13.52	2.736	1.137	11	0.01	106.02	0
106	Vitamin A (KU / kg)	0563	14.03	0.1523	13.52	2.736	1.137	11	0.19	106.02	0
106	Vitamin A (KU / kg)	0910	14.30	0.0000	13.52	2.736	1.137	11	0.29	106.02	0
106	Vitamin A (KU / kg)	0004	14.97	0.4600	13.52	2.736	1.137	11	0.53	106.02	0
106	Vitamin A (KU / kg)	0148	15.06	1.090	13.52	2.736	1.137	11	0.56	106.02	0
106	Vitamin A (KU / kg)	0036	16.25	3.300	13.52	2.736	1.137	11	1.00	106.02	0
106	Vitamin A (KU / kg)	0610	16.45	2.100	13.52	2.736	1.137	11	1.07	106.02	0
106	Vitamin A (KU / kg)	0098	17.20	9.200	13.52	2.736	1.137	11	1.35	106.01	1
106	Vitamin A (KU / kg)	0656	13,367	0.0000	13.52	2.736	1.137	11	4880.18	106.02	2
107	Vitamin B12 (ppb)	0227	36.05	1.500				1		107.00	0
108	Vitamin D3 (KU / kg)	0619	2.630	0.4200	3.676	0.9227	0.4763	8	-1.13	108.99	0
108	Vitamin D3 (KU / kg)	0610	2.950	0.1000	3.676	0.9227	0.4763	8	-0.79	108.02	0
108	Vitamin D3 (KU / kg)	0208	3.070	0.9200	3.676	0.9227	0.4763	8	-0.66	108.02	0
108	Vitamin D3 (KU / kg)	0010	3.460	0.6200	3.676	0.9227	0.4763	8	-0.23	108.99	0
108	Vitamin D3 (KU / kg)	0910	3.740	0.0000	3.676	0.9227	0.4763	8	0.07	108.02	0
108	Vitamin D3 (KU / kg)	0227	4.130	0.2600	3.676	0.9227	0.4763	8	0.49	108.02	0
108	Vitamin D3 (KU / kg)	0870	4.365	0.0700	3.676	0.9227	0.4763	8	0.75	108.99	0
108	Vitamin D3 (KU / kg)	2319	6.460	1.420	3.676	0.9227	0.4763	8	3.02	108.02	0
109	Vitamin E (IU / kg)	2319	32.80	0.0000	52.33	9.708	1.426	11	-2.01	109.02	0
109	Vitamin E (IU / kg)	0619	41.50	1.000	52.33	9.708	1.426	11	-1.12	109.02	0
109	Vitamin E (IU / kg)	0208	45.74	4.008	52.33	9.708	1.426	11	-0.68	109.02	0
109	Vitamin E (IU / kg)	0910	48.50	0.0000	52.33	9.708	1.426	11	-0.39	109.02	0
109	Vitamin E (IU / kg)	0098	52.40	0.0000	52.33	9.708	1.426	11	0.01	109.02	0
109	Vitamin E (IU / kg)	0148	52.68	0.5000	52.33	9.708	1.426	11	0.04	109.02	0
109	Vitamin E (IU / kg)	0610	54.60	2.600	52.33	9.708	1.426	11	0.23	109.02	0
109	Vitamin E (IU / kg)	0227	55.50	1.600	52.33	9.708	1.426	11	0.33	109.02	0
109	Vitamin E (IU / kg)	0036	57.80	2.400	52.33	9.708	1.426	11	0.56	109.02	0
109	Vitamin E (IU / kg)	0905	64.41	0.9100	52.33	9.708	1.426	11	1.24	109.02	0
109	Vitamin E (IU / kg)	0563	64.76	2.669	52.33	9.708	1.426	11	1.28	109.02	0
111	Vitamin C, Phosphorylated (ppm)	0227	< 4.4					0		111.00	5
112	Pyridoxine (µg / g)	0160	3.045	0.2700				3		112.99	0
112	Pyridoxine (µg / g)	0010	4.400	1.200				3		112.01	0
112	Pyridoxine (µg / g)	0227	8.880	0.4600				3		112.01	0
113	Folic Acid (ppm)	0227	6.920	0.6400				1		113.01	0
114	Biotin (ppm)	0010	0.2500	0.1000				2		114.99	0
114	Biotin (ppm)	0227	0.4865	0.0310				2		114.01	0
115	Non Protein N (NPN) (%)	0098	0.5100	0.0600				1		115.00	0
118	Peroxide value (meq/kg)	2081	1.475	0.3100				1		118.99	0
120	Alanine (%)	2246	0.8200	0.0200	0.9475	0.0292	0.0145	29	-4.37	120.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
120	Alanine (%)	2188	0.8660	0.0020	0.9475	0.0292	0.0145	29	-2.79	120.05	0
120	Alanine (%)	0889	0.8875	0.0050	0.9475	0.0292	0.0145	29	-2.06	120.99	0
120	Alanine (%)	0910	0.9050	0.0700	0.9475	0.0292	0.0145	29	-1.46	120.00	0
120	Alanine (%)	0918	0.9181	0.0019	0.9475	0.0292	0.0145	29	-1.01	120.00	0
120	Alanine (%)	0859	0.9199	0.0138	0.9475	0.0292	0.0145	29	-0.95	120.00	0
120	Alanine (%)	0148	0.9250	0.0100	0.9475	0.0292	0.0145	29	-0.77	120.05	0
120	Alanine (%)	0610	0.9300	0.0000	0.9475	0.0292	0.0145	29	-0.60	120.05	0
120	Alanine (%)	0619	0.9385	0.0010	0.9475	0.0292	0.0145	29	-0.31	120.00	0
120	Alanine (%)	0682	0.9390	0.0000	0.9475	0.0292	0.0145	29	-0.29	120.00	0
120	Alanine (%)	0652	0.9400	0.0000	0.9475	0.0292	0.0145	29	-0.26	120.00	0
120	Alanine (%)	0662	0.9437	0.0512	0.9475	0.0292	0.0145	29	-0.13	120.00	0
120	Alanine (%)	0571	0.9470	0.0020	0.9475	0.0292	0.0145	29	-0.02	120.00	0
120	Alanine (%)	0098	0.9475	0.0410	0.9475	0.0292	0.0145	29	0.00	120.02	0
120	Alanine (%)	0407	0.9500	0.0200	0.9475	0.0292	0.0145	29	0.09	120.05	0
120	Alanine (%)	2196	0.9540	0.0000	0.9475	0.0292	0.0145	29	0.22	120.05	0
120	Alanine (%)	0242	0.9550	0.0100	0.9475	0.0292	0.0145	29	0.26	120.05	0
120	Alanine (%)	0644	0.9555	0.0010	0.9475	0.0292	0.0145	29	0.27	120.00	0
120	Alanine (%)	0872	0.9565	0.0010	0.9475	0.0292	0.0145	29	0.31	120.00	0
120	Alanine (%)	0968	0.9600	0.0000	0.9475	0.0292	0.0145	29	0.43	120.00	0
120	Alanine (%)	0878	0.9630	0.0120	0.9475	0.0292	0.0145	29	0.53	120.00	0
120	Alanine (%)	2059	0.9635	0.0030	0.9475	0.0292	0.0145	29	0.55	120.00	0
120	Alanine (%)	0941	0.9695	0.0170	0.9475	0.0292	0.0145	29	0.75	120.00	0
120	Alanine (%)	0870	0.9749	0.0085	0.9475	0.0292	0.0145	29	0.94	120.00	0
120	Alanine (%)	0227	0.9750	0.0100	0.9475	0.0292	0.0145	29	0.94	120.99	0
120	Alanine (%)	0353	0.9795	0.0030	0.9475	0.0292	0.0145	29	1.10	120.00	0
120	Alanine (%)	2144	0.9800	0.0400	0.9475	0.0292	0.0145	29	1.11	120.99	0
120	Alanine (%)	0868	0.9855	0.0130	0.9475	0.0292	0.0145	29	1.30	120.00	0
120	Alanine (%)	0226	1.021	0.0643	0.9475	0.0292	0.0145	29	2.53	120.00	0
121	Arginine (%)	0889	1.013	0.0250	1.294	0.0549	0.0200	28	-5.13	121.99	0
121	Arginine (%)	0859	1.209	0.0215	1.294	0.0549	0.0200	28	-1.54	121.00	0
121	Arginine (%)	2246	1.235	0.0700	1.294	0.0549	0.0200	28	-1.07	121.05	0
121	Arginine (%)	0682	1.235	0.0000	1.294	0.0549	0.0200	28	-1.07	121.00	0
121	Arginine (%)	0918	1.241	0.0040	1.294	0.0549	0.0200	28	-0.96	121.00	0
121	Arginine (%)	0098	1.243	0.0160	1.294	0.0549	0.0200	28	-0.92	121.02	0
121	Arginine (%)	0910	1.250	0.0600	1.294	0.0549	0.0200	28	-0.80	121.00	0
121	Arginine (%)	0870	1.252	0.0201	1.294	0.0549	0.0200	28	-0.76	121.00	0
121	Arginine (%)	0868	1.266	0.0110	1.294	0.0549	0.0200	28	-0.51	121.00	0
121	Arginine (%)	0148	1.270	0.0400	1.294	0.0549	0.0200	28	-0.43	121.05	0
121	Arginine (%)	2196	1.270	0.0000	1.294	0.0549	0.0200	28	-0.43	121.05	0
121	Arginine (%)	0652	1.275	0.0300	1.294	0.0549	0.0200	28	-0.34	121.00	0
121	Arginine (%)	0968	1.285	0.0100	1.294	0.0549	0.0200	28	-0.16	121.00	0
121	Arginine (%)	0619	1.290	0.0000	1.294	0.0549	0.0200	28	-0.07	121.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
121	Arginine (%)	0227	1.290	0.0200	1.294	0.0549	0.0200	28	-0.07	121.99	0
121	Arginine (%)	0644	1.309	0.0180	1.294	0.0549	0.0200	28	0.28	121.00	0
121	Arginine (%)	0662	1.311	0.0027	1.294	0.0549	0.0200	28	0.32	121.00	0
121	Arginine (%)	0571	1.315	0.0100	1.294	0.0549	0.0200	28	0.39	121.00	0
121	Arginine (%)	0878	1.320	0.0230	1.294	0.0549	0.0200	28	0.47	121.00	0
121	Arginine (%)	0407	1.323	0.0030	1.294	0.0549	0.0200	28	0.53	121.05	0
121	Arginine (%)	0941	1.338	0.0160	1.294	0.0549	0.0200	28	0.81	121.00	0
121	Arginine (%)	0242	1.340	0.0200	1.294	0.0549	0.0200	28	0.85	121.05	0
121	Arginine (%)	0610	1.345	0.0100	1.294	0.0549	0.0200	28	0.94	121.05	0
121	Arginine (%)	2059	1.349	0.0050	1.294	0.0549	0.0200	28	1.00	121.00	0
121	Arginine (%)	0872	1.353	0.0230	1.294	0.0549	0.0200	28	1.07	121.00	0
121	Arginine (%)	0226	1.357	0.0401	1.294	0.0549	0.0200	28	1.16	121.00	0
121	Arginine (%)	2188	1.364	0.0110	1.294	0.0549	0.0200	28	1.27	121.05	0
121	Arginine (%)	0353	1.535	0.0500	1.294	0.0549	0.0200	28	4.40	121.00	0
122	Aspartic (%)	0889	1.600	0.0000	1.973	0.0706	0.0213	27	-5.28	122.99	0
122	Aspartic (%)	0859	1.869	0.0336	1.973	0.0706	0.0213	27	-1.46	122.00	0
122	Aspartic (%)	0682	1.876	0.0000	1.973	0.0706	0.0213	27	-1.37	122.00	0
122	Aspartic (%)	0098	1.889	0.0340	1.973	0.0706	0.0213	27	-1.18	122.02	0
122	Aspartic (%)	0918	1.900	0.0350	1.973	0.0706	0.0213	27	-1.04	122.00	0
122	Aspartic (%)	0968	1.920	0.0000	1.973	0.0706	0.0213	27	-0.74	122.00	0
122	Aspartic (%)	2188	1.936	0.0240	1.973	0.0706	0.0213	27	-0.52	122.05	0
122	Aspartic (%)	0571	1.937	0.0080	1.973	0.0706	0.0213	27	-0.50	122.00	0
122	Aspartic (%)	0644	1.944	0.0160	1.973	0.0706	0.0213	27	-0.40	122.00	0
122	Aspartic (%)	0910	1.945	0.0300	1.973	0.0706	0.0213	27	-0.39	122.00	0
122	Aspartic (%)	0652	1.950	0.0200	1.973	0.0706	0.0213	27	-0.32	122.00	0
122	Aspartic (%)	0870	1.968	0.0384	1.973	0.0706	0.0213	27	-0.07	122.00	0
122	Aspartic (%)	0941	1.970	0.0700	1.973	0.0706	0.0213	27	-0.04	122.00	0
122	Aspartic (%)	0619	1.980	0.0200	1.973	0.0706	0.0213	27	0.11	122.00	0
122	Aspartic (%)	0872	1.990	0.0170	1.973	0.0706	0.0213	27	0.24	122.00	0
122	Aspartic (%)	0242	1.995	0.0700	1.973	0.0706	0.0213	27	0.32	122.05	0
122	Aspartic (%)	2196	1.996	0.0000	1.973	0.0706	0.0213	27	0.33	122.05	0
122	Aspartic (%)	2059	1.997	0.0070	1.973	0.0706	0.0213	27	0.34	122.00	0
122	Aspartic (%)	0878	2.000	0.0110	1.973	0.0706	0.0213	27	0.38	122.00	0
122	Aspartic (%)	0610	2.000	0.0000	1.973	0.0706	0.0213	27	0.39	122.05	0
122	Aspartic (%)	0227	2.000	0.0000	1.973	0.0706	0.0213	27	0.39	122.99	0
122	Aspartic (%)	0662	2.006	0.0021	1.973	0.0706	0.0213	27	0.47	122.00	0
122	Aspartic (%)	0868	2.036	0.0140	1.973	0.0706	0.0213	27	0.90	122.00	0
122	Aspartic (%)	0148	2.055	0.0300	1.973	0.0706	0.0213	27	1.17	122.05	0
122	Aspartic (%)	0226	2.120	0.0661	1.973	0.0706	0.0213	27	2.08	122.00	0
122	Aspartic (%)	0353	2.185	0.0100	1.973	0.0706	0.0213	27	3.01	122.00	0
122	Aspartic (%)	2144	2.210	0.0200	1.973	0.0706	0.0213	27	3.36	122.99	0
122	Aspartic (%)	0407	1.848	0.1870	1.973	0.0706	0.0213	27	-1.77	122.05	1

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
124	Cysteine/Cystine (%)	0407	0.1596	0.0172	0.3171	0.0307	0.0087	28	-5.13	124.05	0
124	Cysteine/Cystine (%)	0889	0.2675	0.0050	0.3171	0.0307	0.0087	28	-1.62	124.99	0
124	Cysteine/Cystine (%)	0870	0.2763	0.0019	0.3171	0.0307	0.0087	28	-1.33	124.00	0
124	Cysteine/Cystine (%)	0227	0.2800	0.0000	0.3171	0.0307	0.0087	28	-1.21	124.99	0
124	Cysteine/Cystine (%)	0662	0.2817	0.0241	0.3171	0.0307	0.0087	28	-1.15	124.00	0
124	Cysteine/Cystine (%)	0859	0.2895	0.0059	0.3171	0.0307	0.0087	28	-0.90	124.00	0
124	Cysteine/Cystine (%)	2196	0.2980	0.0000	0.3171	0.0307	0.0087	28	-0.62	124.05	0
124	Cysteine/Cystine (%)	0910	0.3000	0.0200	0.3171	0.0307	0.0087	28	-0.56	124.00	0
124	Cysteine/Cystine (%)	0644	0.3060	0.0120	0.3171	0.0307	0.0087	28	-0.36	124.00	0
124	Cysteine/Cystine (%)	2188	0.3065	0.0050	0.3171	0.0307	0.0087	28	-0.34	124.05	0
124	Cysteine/Cystine (%)	0353	0.3150	0.0100	0.3171	0.0307	0.0087	28	-0.07	124.00	0
124	Cysteine/Cystine (%)	0652	0.3150	0.0100	0.3171	0.0307	0.0087	28	-0.07	124.00	0
124	Cysteine/Cystine (%)	0968	0.3150	0.0100	0.3171	0.0307	0.0087	28	-0.07	124.00	0
124	Cysteine/Cystine (%)	0619	0.3180	0.0020	0.3171	0.0307	0.0087	28	0.03	124.00	0
124	Cysteine/Cystine (%)	2059	0.3180	0.0060	0.3171	0.0307	0.0087	28	0.03	124.00	0
124	Cysteine/Cystine (%)	0098	0.3180	0.0060	0.3171	0.0307	0.0087	28	0.03	124.02	0
124	Cysteine/Cystine (%)	0571	0.3215	0.0010	0.3171	0.0307	0.0087	28	0.14	124.00	0
124	Cysteine/Cystine (%)	0878	0.3220	0.0020	0.3171	0.0307	0.0087	28	0.16	124.00	0
124	Cysteine/Cystine (%)	0872	0.3235	0.0230	0.3171	0.0307	0.0087	28	0.21	124.00	0
124	Cysteine/Cystine (%)	0242	0.3350	0.0100	0.3171	0.0307	0.0087	28	0.58	124.05	0
124	Cysteine/Cystine (%)	0941	0.3360	0.0060	0.3171	0.0307	0.0087	28	0.62	124.00	0
124	Cysteine/Cystine (%)	0868	0.3395	0.0070	0.3171	0.0307	0.0087	28	0.73	124.00	0
124	Cysteine/Cystine (%)	0682	0.3430	0.0000	0.3171	0.0307	0.0087	28	0.85	124.00	0
124	Cysteine/Cystine (%)	0918	0.3500	0.0000	0.3171	0.0307	0.0087	28	1.07	124.00	0
124	Cysteine/Cystine (%)	0610	0.3500	0.0000	0.3171	0.0307	0.0087	28	1.07	124.05	0
124	Cysteine/Cystine (%)	0148	0.3550	0.0320	0.3171	0.0307	0.0087	28	1.24	124.05	0
124	Cysteine/Cystine (%)	0226	0.3606	0.0188	0.3171	0.0307	0.0087	28	1.42	124.00	0
124	Cysteine/Cystine (%)	2144	0.5450	0.0100	0.3171	0.0307	0.0087	28	7.43	124.99	0
124	Cysteine/Cystine (%)	2246	0.7350	0.0500	0.3171	0.0307	0.0087	28	13.62	124.05	1
125	Glutamic (%)	0889	2.733	0.0550	3.502	0.1569	0.0391	29	-4.90	125.99	0
125	Glutamic (%)	0859	3.327	0.0573	3.502	0.1569	0.0391	29	-1.12	125.00	0
125	Glutamic (%)	0098	3.338	0.0030	3.502	0.1569	0.0391	29	-1.05	125.02	0
125	Glutamic (%)	2246	3.345	0.0900	3.502	0.1569	0.0391	29	-1.00	125.05	0
125	Glutamic (%)	0918	3.346	0.0620	3.502	0.1569	0.0391	29	-0.99	125.00	0
125	Glutamic (%)	0682	3.365	0.0000	3.502	0.1569	0.0391	29	-0.87	125.00	0
125	Glutamic (%)	0868	3.378	0.0250	3.502	0.1569	0.0391	29	-0.79	125.00	0
125	Glutamic (%)	0652	3.385	0.0700	3.502	0.1569	0.0391	29	-0.75	125.00	0
125	Glutamic (%)	2188	3.407	0.0180	3.502	0.1569	0.0391	29	-0.61	125.05	0
125	Glutamic (%)	0571	3.429	0.0040	3.502	0.1569	0.0391	29	-0.47	125.00	0
125	Glutamic (%)	0968	3.435	0.0100	3.502	0.1569	0.0391	29	-0.43	125.00	0
125	Glutamic (%)	0407	3.441	0.1130	3.502	0.1569	0.0391	29	-0.39	125.05	0
125	Glutamic (%)	0910	3.475	0.0900	3.502	0.1569	0.0391	29	-0.17	125.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
125	Glutamic (%)	2059	3.488	0.0060	3.502	0.1569	0.0391	29	-0.09	125.00	0
125	Glutamic (%)	0644	3.494	0.0150	3.502	0.1569	0.0391	29	-0.05	125.00	0
125	Glutamic (%)	0242	3.495	0.0100	3.502	0.1569	0.0391	29	-0.05	125.05	0
125	Glutamic (%)	0872	3.497	0.0170	3.502	0.1569	0.0391	29	-0.04	125.00	0
125	Glutamic (%)	2196	3.525	0.0000	3.502	0.1569	0.0391	29	0.15	125.05	0
125	Glutamic (%)	0610	3.535	0.0300	3.502	0.1569	0.0391	29	0.21	125.05	0
125	Glutamic (%)	0941	3.538	0.0680	3.502	0.1569	0.0391	29	0.23	125.00	0
125	Glutamic (%)	0227	3.545	0.0300	3.502	0.1569	0.0391	29	0.27	125.99	0
125	Glutamic (%)	0878	3.592	0.0150	3.502	0.1569	0.0391	29	0.57	125.00	0
125	Glutamic (%)	0619	3.645	0.0100	3.502	0.1569	0.0391	29	0.91	125.00	0
125	Glutamic (%)	0870	3.669	0.0253	3.502	0.1569	0.0391	29	1.06	125.00	0
125	Glutamic (%)	0662	3.684	0.0608	3.502	0.1569	0.0391	29	1.16	125.00	0
125	Glutamic (%)	0226	3.707	0.1093	3.502	0.1569	0.0391	29	1.30	125.00	0
125	Glutamic (%)	0148	3.780	0.0800	3.502	0.1569	0.0391	29	1.77	125.05	0
125	Glutamic (%)	0353	3.950	0.0200	3.502	0.1569	0.0391	29	2.85	125.00	0
125	Glutamic (%)	2144	3.990	0.0400	3.502	0.1569	0.0391	29	3.11	125.99	0
126	Glycine (%)	0682	0.7840	0.0000	0.8357	0.0279	0.0133	27	-1.85	126.00	0
126	Glycine (%)	0859	0.7950	0.0122	0.8357	0.0279	0.0133	27	-1.46	126.00	0
126	Glycine (%)	0407	0.7970	0.0400	0.8357	0.0279	0.0133	27	-1.39	126.05	0
126	Glycine (%)	2246	0.8050	0.0500	0.8357	0.0279	0.0133	27	-1.10	126.05	0
126	Glycine (%)	0918	0.8065	0.0042	0.8357	0.0279	0.0133	27	-1.04	126.00	0
126	Glycine (%)	0910	0.8100	0.0400	0.8357	0.0279	0.0133	27	-0.92	126.00	0
126	Glycine (%)	0652	0.8150	0.0100	0.8357	0.0279	0.0133	27	-0.74	126.00	0
126	Glycine (%)	0148	0.8150	0.0300	0.8357	0.0279	0.0133	27	-0.74	126.05	0
126	Glycine (%)	0870	0.8217	0.0086	0.8357	0.0279	0.0133	27	-0.50	126.00	0
126	Glycine (%)	0098	0.8270	0.0040	0.8357	0.0279	0.0133	27	-0.31	126.02	0
126	Glycine (%)	0968	0.8300	0.0000	0.8357	0.0279	0.0133	27	-0.20	126.00	0
126	Glycine (%)	0571	0.8355	0.0010	0.8357	0.0279	0.0133	27	-0.01	126.00	0
126	Glycine (%)	0872	0.8380	0.0100	0.8357	0.0279	0.0133	27	0.08	126.00	0
126	Glycine (%)	0878	0.8410	0.0160	0.8357	0.0279	0.0133	27	0.19	126.00	0
126	Glycine (%)	2188	0.8415	0.0010	0.8357	0.0279	0.0133	27	0.21	126.05	0
126	Glycine (%)	0644	0.8425	0.0050	0.8357	0.0279	0.0133	27	0.24	126.00	0
126	Glycine (%)	0242	0.8450	0.0100	0.8357	0.0279	0.0133	27	0.33	126.05	0
126	Glycine (%)	0619	0.8465	0.0090	0.8357	0.0279	0.0133	27	0.39	126.00	0
126	Glycine (%)	2196	0.8520	0.0000	0.8357	0.0279	0.0133	27	0.58	126.05	0
126	Glycine (%)	0227	0.8550	0.0100	0.8357	0.0279	0.0133	27	0.69	126.99	0
126	Glycine (%)	2059	0.8555	0.0050	0.8357	0.0279	0.0133	27	0.71	126.00	0
126	Glycine (%)	0868	0.8560	0.0080	0.8357	0.0279	0.0133	27	0.73	126.00	0
126	Glycine (%)	0610	0.8600	0.0000	0.8357	0.0279	0.0133	27	0.87	126.05	0
126	Glycine (%)	2144	0.8600	0.0000	0.8357	0.0279	0.0133	27	0.87	126.99	0
126	Glycine (%)	0941	0.8640	0.0160	0.8357	0.0279	0.0133	27	1.01	126.00	0
126	Glycine (%)	0353	0.8965	0.0110	0.8357	0.0279	0.0133	27	2.18	126.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
126	Glycine (%)	0226	0.9240	0.0582	0.8357	0.0279	0.0133	27	3.16	126.00	0
126	Glycine (%)	0662	0.8635	0.0750	0.8357	0.0279	0.0133	27	1.00	126.00	1
126	Glycine (%)	0889	0.4575	0.0050	0.8357	0.0279	0.0133	27	-13.54	126.99	2
127	Histidine (%)	0889	0.4575	0.0050	0.5164	0.0245	0.0107	28	-2.40	127.99	0
127	Histidine (%)	0619	0.4600	0.0260	0.5164	0.0245	0.0107	28	-2.30	127.00	0
127	Histidine (%)	2188	0.4785	0.0010	0.5164	0.0245	0.0107	28	-1.54	127.05	0
127	Histidine (%)	0918	0.4899	0.0085	0.5164	0.0245	0.0107	28	-1.08	127.00	0
127	Histidine (%)	0148	0.4900	0.0200	0.5164	0.0245	0.0107	28	-1.07	127.05	0
127	Histidine (%)	0910	0.4950	0.0100	0.5164	0.0245	0.0107	28	-0.87	127.00	0
127	Histidine (%)	2246	0.4950	0.0300	0.5164	0.0245	0.0107	28	-0.87	127.05	0
127	Histidine (%)	0098	0.5005	0.0110	0.5164	0.0245	0.0107	28	-0.65	127.02	0
127	Histidine (%)	0968	0.5100	0.0000	0.5164	0.0245	0.0107	28	-0.26	127.00	0
127	Histidine (%)	0859	0.5127	0.0077	0.5164	0.0245	0.0107	28	-0.15	127.00	0
127	Histidine (%)	0682	0.5150	0.0000	0.5164	0.0245	0.0107	28	-0.06	127.00	0
127	Histidine (%)	2059	0.5155	0.0070	0.5164	0.0245	0.0107	28	-0.03	127.00	0
127	Histidine (%)	0870	0.5159	0.0303	0.5164	0.0245	0.0107	28	-0.02	127.00	0
127	Histidine (%)	0644	0.5175	0.0030	0.5164	0.0245	0.0107	28	0.05	127.00	0
127	Histidine (%)	0571	0.5180	0.0040	0.5164	0.0245	0.0107	28	0.07	127.00	0
127	Histidine (%)	0407	0.5185	0.0030	0.5164	0.0245	0.0107	28	0.09	127.05	0
127	Histidine (%)	0242	0.5200	0.0400	0.5164	0.0245	0.0107	28	0.15	127.05	0
127	Histidine (%)	0610	0.5200	0.0000	0.5164	0.0245	0.0107	28	0.15	127.05	0
127	Histidine (%)	0662	0.5206	0.0057	0.5164	0.0245	0.0107	28	0.17	127.00	0
127	Histidine (%)	0868	0.5295	0.0030	0.5164	0.0245	0.0107	28	0.54	127.00	0
127	Histidine (%)	0878	0.5325	0.0030	0.5164	0.0245	0.0107	28	0.66	127.00	0
127	Histidine (%)	0941	0.5335	0.0170	0.5164	0.0245	0.0107	28	0.70	127.00	0
127	Histidine (%)	0872	0.5345	0.0150	0.5164	0.0245	0.0107	28	0.74	127.00	0
127	Histidine (%)	0652	0.5400	0.0000	0.5164	0.0245	0.0107	28	0.96	127.00	0
127	Histidine (%)	2196	0.5440	0.0000	0.5164	0.0245	0.0107	28	1.13	127.05	0
127	Histidine (%)	0353	0.5450	0.0100	0.5164	0.0245	0.0107	28	1.17	127.00	0
127	Histidine (%)	0227	0.5550	0.0300	0.5164	0.0245	0.0107	28	1.57	127.99	0
127	Histidine (%)	2144	0.7850	0.0100	0.5164	0.0245	0.0107	28	10.95	127.99	0
127	Histidine (%)	0226	0.6023	0.1637	0.5164	0.0245	0.0107	28	3.50	127.00	1
128	Isoleucine (%)	0889	0.7175	0.0050	0.7896	0.0504	0.0159	28	-1.43	128.99	0
128	Isoleucine (%)	0242	0.7200	0.1200	0.7896	0.0504	0.0159	28	-1.38	128.05	0
128	Isoleucine (%)	0918	0.7201	0.0108	0.7896	0.0504	0.0159	28	-1.38	128.00	0
128	Isoleucine (%)	0870	0.7207	0.0235	0.7896	0.0504	0.0159	28	-1.37	128.00	0
128	Isoleucine (%)	0859	0.7348	0.0131	0.7896	0.0504	0.0159	28	-1.09	128.00	0
128	Isoleucine (%)	0682	0.7470	0.0000	0.7896	0.0504	0.0159	28	-0.85	128.00	0
128	Isoleucine (%)	2188	0.7535	0.0010	0.7896	0.0504	0.0159	28	-0.72	128.05	0
128	Isoleucine (%)	2246	0.7550	0.0300	0.7896	0.0504	0.0159	28	-0.69	128.05	0
128	Isoleucine (%)	0353	0.7650	0.0100	0.7896	0.0504	0.0159	28	-0.49	128.00	0
128	Isoleucine (%)	0872	0.7755	0.0010	0.7896	0.0504	0.0159	28	-0.28	128.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
128	Isoleucine (%)	0652	0.7800	0.0000	0.7896	0.0504	0.0159	28	-0.19	128.00	0
128	Isoleucine (%)	0968	0.7800	0.0000	0.7896	0.0504	0.0159	28	-0.19	128.00	0
128	Isoleucine (%)	0868	0.7820	0.0060	0.7896	0.0504	0.0159	28	-0.15	128.00	0
128	Isoleucine (%)	0910	0.7900	0.0400	0.7896	0.0504	0.0159	28	0.01	128.00	0
128	Isoleucine (%)	0878	0.7940	0.0240	0.7896	0.0504	0.0159	28	0.09	128.00	0
128	Isoleucine (%)	0227	0.8000	0.0200	0.7896	0.0504	0.0159	28	0.21	128.99	0
128	Isoleucine (%)	2196	0.8010	0.0000	0.7896	0.0504	0.0159	28	0.23	128.05	0
128	Isoleucine (%)	0571	0.8055	0.0070	0.7896	0.0504	0.0159	28	0.32	128.00	0
128	Isoleucine (%)	0098	0.8055	0.0030	0.7896	0.0504	0.0159	28	0.32	128.02	0
128	Isoleucine (%)	0619	0.8160	0.0300	0.7896	0.0504	0.0159	28	0.52	128.00	0
128	Isoleucine (%)	0941	0.8295	0.0110	0.7896	0.0504	0.0159	28	0.79	128.00	0
128	Isoleucine (%)	0644	0.8355	0.0110	0.7896	0.0504	0.0159	28	0.91	128.00	0
128	Isoleucine (%)	0662	0.8369	0.0184	0.7896	0.0504	0.0159	28	0.94	128.00	0
128	Isoleucine (%)	2059	0.8415	0.0050	0.7896	0.0504	0.0159	28	1.03	128.00	0
128	Isoleucine (%)	0407	0.8420	0.0260	0.7896	0.0504	0.0159	28	1.04	128.05	0
128	Isoleucine (%)	0610	0.8450	0.0100	0.7896	0.0504	0.0159	28	1.10	128.05	0
128	Isoleucine (%)	0148	0.8500	0.0200	0.7896	0.0504	0.0159	28	1.20	128.05	0
128	Isoleucine (%)	2144	0.9000	0.0000	0.7896	0.0504	0.0159	28	2.19	128.99	0
128	Isoleucine (%)	0226	0.9196	0.1683	0.7896	0.0504	0.0159	28	2.58	128.00	1
129	Leucine (%)	0682	1.423	0.0000	1.545	0.0494	0.0161	28	-2.48	129.00	0
129	Leucine (%)	0859	1.455	0.0241	1.545	0.0494	0.0161	28	-1.82	129.00	0
129	Leucine (%)	0652	1.485	0.0100	1.545	0.0494	0.0161	28	-1.22	129.00	0
129	Leucine (%)	2188	1.486	0.0050	1.545	0.0494	0.0161	28	-1.21	129.05	0
129	Leucine (%)	2246	1.490	0.0600	1.545	0.0494	0.0161	28	-1.12	129.05	0
129	Leucine (%)	0662	1.495	0.0019	1.545	0.0494	0.0161	28	-1.02	129.00	0
129	Leucine (%)	0910	1.500	0.0400	1.545	0.0494	0.0161	28	-0.92	129.00	0
129	Leucine (%)	2196	1.520	0.0000	1.545	0.0494	0.0161	28	-0.52	129.05	0
129	Leucine (%)	0242	1.525	0.0500	1.545	0.0494	0.0161	28	-0.41	129.05	0
129	Leucine (%)	0098	1.532	0.0090	1.545	0.0494	0.0161	28	-0.28	129.02	0
129	Leucine (%)	0870	1.539	0.0582	1.545	0.0494	0.0161	28	-0.14	129.00	0
129	Leucine (%)	0148	1.540	0.0200	1.545	0.0494	0.0161	28	-0.11	129.05	0
129	Leucine (%)	0407	1.545	0.0330	1.545	0.0494	0.0161	28	-0.02	129.05	0
129	Leucine (%)	2144	1.545	0.0100	1.545	0.0494	0.0161	28	-0.01	129.99	0
129	Leucine (%)	0918	1.552	0.0145	1.545	0.0494	0.0161	28	0.13	129.00	0
129	Leucine (%)	0619	1.555	0.0100	1.545	0.0494	0.0161	28	0.19	129.00	0
129	Leucine (%)	0968	1.555	0.0100	1.545	0.0494	0.0161	28	0.19	129.00	0
129	Leucine (%)	0571	1.556	0.0010	1.545	0.0494	0.0161	28	0.20	129.00	0
129	Leucine (%)	0644	1.572	0.0040	1.545	0.0494	0.0161	28	0.54	129.00	0
129	Leucine (%)	0878	1.574	0.0000	1.545	0.0494	0.0161	28	0.58	129.00	0
129	Leucine (%)	0872	1.578	0.0130	1.545	0.0494	0.0161	28	0.65	129.00	0
129	Leucine (%)	2059	1.581	0.0010	1.545	0.0494	0.0161	28	0.71	129.00	0
129	Leucine (%)	0610	1.585	0.0100	1.545	0.0494	0.0161	28	0.80	129.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
129	Leucine (%)	0941	1.589	0.0330	1.545	0.0494	0.0161	28	0.87	129.00	0
129	Leucine (%)	0227	1.595	0.0100	1.545	0.0494	0.0161	28	1.00	129.99	0
129	Leucine (%)	0868	1.601	0.0090	1.545	0.0494	0.0161	28	1.11	129.00	0
129	Leucine (%)	0889	1.723	0.0050	1.545	0.0494	0.0161	28	3.59	129.99	0
129	Leucine (%)	0353	1.745	0.0100	1.545	0.0494	0.0161	28	4.04	129.00	0
129	Leucine (%)	0226	1.716	0.2185	1.545	0.0494	0.0161	28	3.45	129.00	1
130	L-Lysine (%)	0889	0.6200	0.0000	1.047	0.0580	0.0177	30	-7.37	130.99	0
130	L-Lysine (%)	0407	0.9260	0.0320	1.047	0.0580	0.0177	30	-2.09	130.05	0
130	L-Lysine (%)	0682	0.9690	0.0000	1.047	0.0580	0.0177	30	-1.35	130.00	0
130	L-Lysine (%)	0859	0.9807	0.0169	1.047	0.0580	0.0177	30	-1.15	130.00	0
130	L-Lysine (%)	0652	0.9900	0.0000	1.047	0.0580	0.0177	30	-0.99	130.00	0
130	L-Lysine (%)	2246	0.9900	0.0000	1.047	0.0580	0.0177	30	-0.99	130.05	0
130	L-Lysine (%)	0910	0.9950	0.0100	1.047	0.0580	0.0177	30	-0.90	130.00	0
130	L-Lysine (%)	0662	1.022	0.0136	1.047	0.0580	0.0177	30	-0.45	130.00	0
130	L-Lysine (%)	0918	1.030	0.0301	1.047	0.0580	0.0177	30	-0.31	130.00	0
130	L-Lysine (%)	0571	1.031	0.0040	1.047	0.0580	0.0177	30	-0.28	130.00	0
130	L-Lysine (%)	0872	1.037	0.0320	1.047	0.0580	0.0177	30	-0.18	130.00	0
130	L-Lysine (%)	0644	1.039	0.0040	1.047	0.0580	0.0177	30	-0.15	130.00	0
130	L-Lysine (%)	0610	1.040	0.0200	1.047	0.0580	0.0177	30	-0.13	130.05	0
130	L-Lysine (%)	2196	1.040	0.0000	1.047	0.0580	0.0177	30	-0.13	130.05	0
130	L-Lysine (%)	0619	1.045	0.0100	1.047	0.0580	0.0177	30	-0.04	130.00	0
130	L-Lysine (%)	0878	1.047	0.0000	1.047	0.0580	0.0177	30	-0.01	130.00	0
130	L-Lysine (%)	0868	1.048	0.0160	1.047	0.0580	0.0177	30	0.01	130.00	0
130	L-Lysine (%)	0941	1.050	0.0150	1.047	0.0580	0.0177	30	0.04	130.00	0
130	L-Lysine (%)	2059	1.055	0.0090	1.047	0.0580	0.0177	30	0.12	130.00	0
130	L-Lysine (%)	0014	1.060	0.0400	1.047	0.0580	0.0177	30	0.22	130.05	0
130	L-Lysine (%)	0242	1.060	0.0400	1.047	0.0580	0.0177	30	0.22	130.05	0
130	L-Lysine (%)	0870	1.080	0.0343	1.047	0.0580	0.0177	30	0.56	130.00	0
130	L-Lysine (%)	0968	1.080	0.0000	1.047	0.0580	0.0177	30	0.56	130.00	0
130	L-Lysine (%)	0148	1.080	0.0400	1.047	0.0580	0.0177	30	0.56	130.05	0
130	L-Lysine (%)	0098	1.088	0.0430	1.047	0.0580	0.0177	30	0.69	130.02	0
130	L-Lysine (%)	0227	1.115	0.0300	1.047	0.0580	0.0177	30	1.16	130.99	0
130	L-Lysine (%)	0353	1.130	0.0200	1.047	0.0580	0.0177	30	1.42	130.00	0
130	L-Lysine (%)	2188	1.162	0.0010	1.047	0.0580	0.0177	30	1.97	130.05	0
130	L-Lysine (%)	0226	1.163	0.0612	1.047	0.0580	0.0177	30	1.99	130.00	0
130	L-Lysine (%)	2144	1.305	0.0100	1.047	0.0580	0.0177	30	4.44	130.99	0
130	L-Lysine (%)	0027	1.248	0.2860	1.047	0.0580	0.0177	30	3.46	130.05	1
131	Methionine (%)	0407	0.1256	0.0091	0.2919	0.0232	0.0073	29	-7.17	131.05	0
131	Methionine (%)	0148	0.2550	0.0100	0.2919	0.0232	0.0073	29	-1.59	131.05	0
131	Methionine (%)	0682	0.2570	0.0000	0.2919	0.0232	0.0073	29	-1.51	131.00	0
131	Methionine (%)	0941	0.2700	0.0100	0.2919	0.0232	0.0073	29	-0.94	131.00	0
131	Methionine (%)	2246	0.2700	0.0000	0.2919	0.0232	0.0073	29	-0.94	131.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
131	Methionine (%)	0859	0.2734	0.0046	0.2919	0.0232	0.0073	29	-0.80	131.00	0
131	Methionine (%)	0918	0.2750	0.0100	0.2919	0.0232	0.0073	29	-0.73	131.00	0
131	Methionine (%)	0242	0.2750	0.0100	0.2919	0.0232	0.0073	29	-0.73	131.05	0
131	Methionine (%)	2188	0.2755	0.0050	0.2919	0.0232	0.0073	29	-0.71	131.05	0
131	Methionine (%)	0652	0.2800	0.0000	0.2919	0.0232	0.0073	29	-0.51	131.00	0
131	Methionine (%)	0098	0.2810	0.0000	0.2919	0.0232	0.0073	29	-0.47	131.02	0
131	Methionine (%)	0968	0.2850	0.0100	0.2919	0.0232	0.0073	29	-0.30	131.00	0
131	Methionine (%)	0868	0.2930	0.0020	0.2919	0.0232	0.0073	29	0.05	131.00	0
131	Methionine (%)	0571	0.2935	0.0050	0.2919	0.0232	0.0073	29	0.07	131.00	0
131	Methionine (%)	0910	0.2950	0.0100	0.2919	0.0232	0.0073	29	0.13	131.00	0
131	Methionine (%)	0227	0.2950	0.0100	0.2919	0.0232	0.0073	29	0.13	131.99	0
131	Methionine (%)	0662	0.2983	0.0147	0.2919	0.0232	0.0073	29	0.27	131.00	0
131	Methionine (%)	0872	0.2990	0.0000	0.2919	0.0232	0.0073	29	0.31	131.00	0
131	Methionine (%)	0610	0.3000	0.0000	0.2919	0.0232	0.0073	29	0.35	131.05	0
131	Methionine (%)	0619	0.3035	0.0130	0.2919	0.0232	0.0073	29	0.50	131.00	0
131	Methionine (%)	2059	0.3050	0.0100	0.2919	0.0232	0.0073	29	0.56	131.00	0
131	Methionine (%)	0878	0.3055	0.0030	0.2919	0.0232	0.0073	29	0.59	131.00	0
131	Methionine (%)	0870	0.3084	0.0139	0.2919	0.0232	0.0073	29	0.71	131.00	0
131	Methionine (%)	2196	0.3090	0.0000	0.2919	0.0232	0.0073	29	0.74	131.05	0
131	Methionine (%)	0226	0.3119	0.0238	0.2919	0.0232	0.0073	29	0.86	131.00	0
131	Methionine (%)	0644	0.3140	0.0180	0.2919	0.0232	0.0073	29	0.95	131.00	0
131	Methionine (%)	0889	0.3250	0.0000	0.2919	0.0232	0.0073	29	1.43	131.99	0
131	Methionine (%)	0353	0.3350	0.0100	0.2919	0.0232	0.0073	29	1.86	131.00	0
131	Methionine (%)	2144	0.3350	0.0100	0.2919	0.0232	0.0073	29	1.86	131.99	0
132	Phenylalanine (%)	0682	0.8150	0.0000	0.9268	0.0404	0.0132	28	-2.77	132.00	0
132	Phenylalanine (%)	0619	0.8620	0.0200	0.9268	0.0404	0.0132	28	-1.61	132.00	0
132	Phenylalanine (%)	0889	0.8700	0.0000	0.9268	0.0404	0.0132	28	-1.41	132.99	0
132	Phenylalanine (%)	0910	0.8800	0.0400	0.9268	0.0404	0.0132	28	-1.16	132.00	0
132	Phenylalanine (%)	0242	0.8900	0.0000	0.9268	0.0404	0.0132	28	-0.91	132.05	0
132	Phenylalanine (%)	0148	0.8950	0.0300	0.9268	0.0404	0.0132	28	-0.79	132.05	0
132	Phenylalanine (%)	0652	0.9000	0.0000	0.9268	0.0404	0.0132	28	-0.66	132.00	0
132	Phenylalanine (%)	2246	0.9000	0.0600	0.9268	0.0404	0.0132	28	-0.66	132.05	0
132	Phenylalanine (%)	0868	0.9015	0.0150	0.9268	0.0404	0.0132	28	-0.63	132.00	0
132	Phenylalanine (%)	0098	0.9050	0.0120	0.9268	0.0404	0.0132	28	-0.54	132.02	0
132	Phenylalanine (%)	0870	0.9201	0.0176	0.9268	0.0404	0.0132	28	-0.17	132.00	0
132	Phenylalanine (%)	0859	0.9212	0.0264	0.9268	0.0404	0.0132	28	-0.14	132.00	0
132	Phenylalanine (%)	0878	0.9270	0.0100	0.9268	0.0404	0.0132	28	0.00	132.00	0
132	Phenylalanine (%)	2144	0.9300	0.0200	0.9268	0.0404	0.0132	28	0.08	132.99	0
132	Phenylalanine (%)	0968	0.9350	0.0100	0.9268	0.0404	0.0132	28	0.20	132.00	0
132	Phenylalanine (%)	0407	0.9355	0.0170	0.9268	0.0404	0.0132	28	0.21	132.05	0
132	Phenylalanine (%)	0662	0.9381	0.0087	0.9268	0.0404	0.0132	28	0.28	132.00	0
132	Phenylalanine (%)	0571	0.9385	0.0010	0.9268	0.0404	0.0132	28	0.29	132.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
132	Phenylalanine (%)	2196	0.9440	0.0000	0.9268	0.0404	0.0132	28	0.42	132.05	0
132	Phenylalanine (%)	0227	0.9450	0.0100	0.9268	0.0404	0.0132	28	0.45	132.99	0
132	Phenylalanine (%)	0644	0.9500	0.0060	0.9268	0.0404	0.0132	28	0.57	132.00	0
132	Phenylalanine (%)	2188	0.9535	0.0010	0.9268	0.0404	0.0132	28	0.66	132.05	0
132	Phenylalanine (%)	0610	0.9550	0.0100	0.9268	0.0404	0.0132	28	0.70	132.05	0
132	Phenylalanine (%)	0872	0.9585	0.0250	0.9268	0.0404	0.0132	28	0.78	132.00	0
132	Phenylalanine (%)	2059	0.9640	0.0080	0.9268	0.0404	0.0132	28	0.92	132.00	0
132	Phenylalanine (%)	0941	1.004	0.0010	0.9268	0.0404	0.0132	28	1.90	132.00	0
132	Phenylalanine (%)	0918	1.080	0.0113	0.9268	0.0404	0.0132	28	3.78	132.00	0
132	Phenylalanine (%)	0353	1.095	0.0100	0.9268	0.0404	0.0132	28	4.16	132.00	0
132	Phenylalanine (%)	0226	1.042	0.0735	0.9268	0.0404	0.0132	28	2.86	132.00	1
133	Proline (%)	0859	0.8293	0.0161	1.154	0.0762	0.0270	27	-4.25	133.00	0
133	Proline (%)	0889	1.033	0.0050	1.154	0.0762	0.0270	27	-1.59	133.99	0
133	Proline (%)	0619	1.065	0.0500	1.154	0.0762	0.0270	27	-1.16	133.00	0
133	Proline (%)	0652	1.075	0.0500	1.154	0.0762	0.0270	27	-1.03	133.00	0
133	Proline (%)	0870	1.085	0.0016	1.154	0.0762	0.0270	27	-0.90	133.00	0
133	Proline (%)	0910	1.090	0.0800	1.154	0.0762	0.0270	27	-0.83	133.00	0
133	Proline (%)	2246	1.115	0.0500	1.154	0.0762	0.0270	27	-0.51	133.05	0
133	Proline (%)	0571	1.119	0.0250	1.154	0.0762	0.0270	27	-0.46	133.00	0
133	Proline (%)	0644	1.125	0.0050	1.154	0.0762	0.0270	27	-0.38	133.00	0
133	Proline (%)	0872	1.126	0.0480	1.154	0.0762	0.0270	27	-0.36	133.00	0
133	Proline (%)	0407	1.136	0.0270	1.154	0.0762	0.0270	27	-0.24	133.05	0
133	Proline (%)	0878	1.144	0.0000	1.154	0.0762	0.0270	27	-0.13	133.00	0
133	Proline (%)	2059	1.149	0.0020	1.154	0.0762	0.0270	27	-0.06	133.00	0
133	Proline (%)	0148	1.150	0.0000	1.154	0.0762	0.0270	27	-0.05	133.05	0
133	Proline (%)	0610	1.150	0.0000	1.154	0.0762	0.0270	27	-0.05	133.05	0
133	Proline (%)	0868	1.155	0.0030	1.154	0.0762	0.0270	27	0.01	133.00	0
133	Proline (%)	0227	1.160	0.0400	1.154	0.0762	0.0270	27	0.08	133.99	0
133	Proline (%)	0242	1.165	0.0500	1.154	0.0762	0.0270	27	0.15	133.05	0
133	Proline (%)	0968	1.175	0.0100	1.154	0.0762	0.0270	27	0.28	133.00	0
133	Proline (%)	0941	1.180	0.0040	1.154	0.0762	0.0270	27	0.35	133.00	0
133	Proline (%)	0682	1.206	0.0000	1.154	0.0762	0.0270	27	0.69	133.00	0
133	Proline (%)	2188	1.213	0.0070	1.154	0.0762	0.0270	27	0.77	133.05	0
133	Proline (%)	2196	1.216	0.0000	1.154	0.0762	0.0270	27	0.82	133.05	0
133	Proline (%)	2144	1.285	0.0100	1.154	0.0762	0.0270	27	1.72	133.99	0
133	Proline (%)	0353	1.310	0.1200	1.154	0.0762	0.0270	27	2.05	133.00	0
133	Proline (%)	0662	1.317	0.0156	1.154	0.0762	0.0270	27	2.14	133.00	0
133	Proline (%)	0918	1.382	0.1097	1.154	0.0762	0.0270	27	2.99	133.00	0
133	Proline (%)	0226	1.169	0.2561	1.154	0.0762	0.0270	27	0.21	133.00	1
134	Serine (%)	0889	0.7200	0.0000	0.9381	0.0505	0.0140	28	-4.32	134.99	0
134	Serine (%)	0407	0.7220	0.0240	0.9381	0.0505	0.0140	28	-4.28	134.05	0
134	Serine (%)	0098	0.7515	0.0370	0.9381	0.0505	0.0140	28	-3.69	134.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
134	Serine (%)	0610	0.8800	0.0000	0.9381	0.0505	0.0140	28	-1.15	134.05	0
134	Serine (%)	0148	0.8950	0.0300	0.9381	0.0505	0.0140	28	-0.85	134.05	0
134	Serine (%)	0644	0.9105	0.0090	0.9381	0.0505	0.0140	28	-0.55	134.00	0
134	Serine (%)	0910	0.9150	0.0300	0.9381	0.0505	0.0140	28	-0.46	134.00	0
134	Serine (%)	2059	0.9155	0.0010	0.9381	0.0505	0.0140	28	-0.45	134.00	0
134	Serine (%)	0571	0.9190	0.0060	0.9381	0.0505	0.0140	28	-0.38	134.00	0
134	Serine (%)	2188	0.9190	0.0100	0.9381	0.0505	0.0140	28	-0.38	134.05	0
134	Serine (%)	0682	0.9220	0.0000	0.9381	0.0505	0.0140	28	-0.32	134.00	0
134	Serine (%)	0868	0.9225	0.0090	0.9381	0.0505	0.0140	28	-0.31	134.00	0
134	Serine (%)	0652	0.9250	0.0100	0.9381	0.0505	0.0140	28	-0.26	134.00	0
134	Serine (%)	0662	0.9294	0.0197	0.9381	0.0505	0.0140	28	-0.17	134.00	0
134	Serine (%)	0859	0.9301	0.0153	0.9381	0.0505	0.0140	28	-0.16	134.00	0
134	Serine (%)	0968	0.9350	0.0100	0.9381	0.0505	0.0140	28	-0.06	134.00	0
134	Serine (%)	0242	0.9550	0.0300	0.9381	0.0505	0.0140	28	0.34	134.05	0
134	Serine (%)	0619	0.9565	0.0050	0.9381	0.0505	0.0140	28	0.36	134.00	0
134	Serine (%)	0872	0.9570	0.0060	0.9381	0.0505	0.0140	28	0.37	134.00	0
134	Serine (%)	0878	0.9595	0.0050	0.9381	0.0505	0.0140	28	0.42	134.00	0
134	Serine (%)	0918	0.9636	0.0174	0.9381	0.0505	0.0140	28	0.51	134.00	0
134	Serine (%)	2246	0.9650	0.0500	0.9381	0.0505	0.0140	28	0.53	134.05	0
134	Serine (%)	0941	0.9740	0.0140	0.9381	0.0505	0.0140	28	0.71	134.00	0
134	Serine (%)	0227	0.9850	0.0100	0.9381	0.0505	0.0140	28	0.93	134.99	0
134	Serine (%)	2196	1.004	0.0000	0.9381	0.0505	0.0140	28	1.31	134.05	0
134	Serine (%)	0870	1.030	0.0146	0.9381	0.0505	0.0140	28	1.82	134.00	0
134	Serine (%)	2144	1.035	0.0100	0.9381	0.0505	0.0140	28	1.92	134.99	0
134	Serine (%)	0353	1.150	0.0200	0.9381	0.0505	0.0140	28	4.20	134.00	0
134	Serine (%)	0226	1.065	0.1020	0.9381	0.0505	0.0140	28	2.50	134.00	1
135	Threonine (%)	0407	0.5870	0.0220	0.7272	0.0349	0.0081	28	-4.02	135.05	0
135	Threonine (%)	0889	0.6100	0.0000	0.7272	0.0349	0.0081	28	-3.36	135.99	0
135	Threonine (%)	0682	0.6760	0.0000	0.7272	0.0349	0.0081	28	-1.47	135.00	0
135	Threonine (%)	2196	0.6900	0.0000	0.7272	0.0349	0.0081	28	-1.07	135.05	0
135	Threonine (%)	0652	0.6950	0.0100	0.7272	0.0349	0.0081	28	-0.92	135.00	0
135	Threonine (%)	0148	0.6950	0.0100	0.7272	0.0349	0.0081	28	-0.92	135.05	0
135	Threonine (%)	0098	0.6955	0.0090	0.7272	0.0349	0.0081	28	-0.91	135.02	0
135	Threonine (%)	0859	0.6981	0.0116	0.7272	0.0349	0.0081	28	-0.83	135.00	0
135	Threonine (%)	0662	0.7027	0.0068	0.7272	0.0349	0.0081	28	-0.70	135.00	0
135	Threonine (%)	0910	0.7150	0.0100	0.7272	0.0349	0.0081	28	-0.35	135.00	0
135	Threonine (%)	0868	0.7165	0.0070	0.7272	0.0349	0.0081	28	-0.31	135.00	0
135	Threonine (%)	0918	0.7187	0.0015	0.7272	0.0349	0.0081	28	-0.25	135.00	0
135	Threonine (%)	0610	0.7300	0.0000	0.7272	0.0349	0.0081	28	0.08	135.05	0
135	Threonine (%)	0644	0.7345	0.0090	0.7272	0.0349	0.0081	28	0.21	135.00	0
135	Threonine (%)	0872	0.7355	0.0010	0.7272	0.0349	0.0081	28	0.24	135.00	0
135	Threonine (%)	0242	0.7400	0.0400	0.7272	0.0349	0.0081	28	0.37	135.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
135	Threonine (%)	2059	0.7405	0.0030	0.7272	0.0349	0.0081	28	0.38	135.00	0
135	Threonine (%)	0941	0.7460	0.0120	0.7272	0.0349	0.0081	28	0.54	135.00	0
135	Threonine (%)	0571	0.7470	0.0020	0.7272	0.0349	0.0081	28	0.57	135.00	0
135	Threonine (%)	0968	0.7500	0.0000	0.7272	0.0349	0.0081	28	0.65	135.00	0
135	Threonine (%)	0878	0.7530	0.0020	0.7272	0.0349	0.0081	28	0.74	135.00	0
135	Threonine (%)	0227	0.7550	0.0100	0.7272	0.0349	0.0081	28	0.80	135.99	0
135	Threonine (%)	0870	0.7587	0.0038	0.7272	0.0349	0.0081	28	0.90	135.00	0
135	Threonine (%)	2188	0.7595	0.0050	0.7272	0.0349	0.0081	28	0.93	135.05	0
135	Threonine (%)	0619	0.7600	0.0020	0.7272	0.0349	0.0081	28	0.94	135.00	0
135	Threonine (%)	2246	0.7600	0.0400	0.7272	0.0349	0.0081	28	0.94	135.05	0
135	Threonine (%)	2144	0.7600	0.0000	0.7272	0.0349	0.0081	28	0.94	135.99	0
135	Threonine (%)	0353	0.8350	0.0100	0.7272	0.0349	0.0081	28	3.09	135.00	0
135	Threonine (%)	0226	0.8301	0.1685	0.7272	0.0349	0.0081	28	2.95	135.00	1
136	Tryptophan (%)	2246	0.1550	0.0100	0.2564	0.0302	0.0068	22	-3.35	136.05	0
136	Tryptophan (%)	0910	0.2100	0.0000	0.2564	0.0302	0.0068	22	-1.54	136.00	0
136	Tryptophan (%)	0662	0.2273	0.0056	0.2564	0.0302	0.0068	22	-0.96	136.00	0
136	Tryptophan (%)	0941	0.2320	0.0320	0.2564	0.0302	0.0068	22	-0.81	136.00	0
136	Tryptophan (%)	0098	0.2385	0.0050	0.2564	0.0302	0.0068	22	-0.59	136.02	0
136	Tryptophan (%)	0682	0.2390	0.0000	0.2564	0.0302	0.0068	22	-0.58	136.00	0
136	Tryptophan (%)	0148	0.2450	0.0020	0.2564	0.0302	0.0068	22	-0.38	136.03	0
136	Tryptophan (%)	0859	0.2453	0.0145	0.2564	0.0302	0.0068	22	-0.37	136.03	0
136	Tryptophan (%)	2059	0.2500	0.0000	0.2564	0.0302	0.0068	22	-0.21	136.03	0
136	Tryptophan (%)	2319	0.2500	0.0000	0.2564	0.0302	0.0068	22	-0.21	136.05	0
136	Tryptophan (%)	0644	0.2515	0.0010	0.2564	0.0302	0.0068	22	-0.16	136.01	0
136	Tryptophan (%)	0571	0.2535	0.0090	0.2564	0.0302	0.0068	22	-0.10	136.03	0
136	Tryptophan (%)	0872	0.2540	0.0020	0.2564	0.0302	0.0068	22	-0.08	136.03	0
136	Tryptophan (%)	0878	0.2575	0.0030	0.2564	0.0302	0.0068	22	0.04	136.01	0
136	Tryptophan (%)	0619	0.2580	0.0120	0.2564	0.0302	0.0068	22	0.05	136.00	0
136	Tryptophan (%)	0868	0.2600	0.0000	0.2564	0.0302	0.0068	22	0.12	136.01	0
136	Tryptophan (%)	0918	0.2850	0.0300	0.2564	0.0302	0.0068	22	0.95	136.00	0
136	Tryptophan (%)	0353	0.2850	0.0100	0.2564	0.0302	0.0068	22	0.95	136.03	0
136	Tryptophan (%)	0242	0.2850	0.0100	0.2564	0.0302	0.0068	22	0.95	136.05	0
136	Tryptophan (%)	0227	0.3000	0.0000	0.2564	0.0302	0.0068	22	1.44	136.00	0
136	Tryptophan (%)	0870	0.3005	0.0032	0.2564	0.0302	0.0068	22	1.46	136.00	0
136	Tryptophan (%)	0889	0.5000	0.0000	0.2564	0.0302	0.0068	22	8.06	136.99	0
136	Tryptophan (%)	0407	0.6315	0.0550	0.2564	0.0302	0.0068	22	12.41	136.05	1
137	Tyrosine (%)	0910	0.3700	0.0000	0.6399	0.1010	0.0156	22	-2.67	137.00	0
137	Tyrosine (%)	2246	0.4700	0.0200	0.6399	0.1010	0.0156	22	-1.68	137.05	0
137	Tyrosine (%)	0889	0.5000	0.0000	0.6399	0.1010	0.0156	22	-1.39	137.99	0
137	Tyrosine (%)	0968	0.5550	0.0100	0.6399	0.1010	0.0156	22	-0.84	137.00	0
137	Tyrosine (%)	0870	0.5643	0.0038	0.6399	0.1010	0.0156	22	-0.75	137.00	0
137	Tyrosine (%)	0098	0.5885	0.0190	0.6399	0.1010	0.0156	22	-0.51	137.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
137	Tyrosine (%)	0242	0.5900	0.0600	0.6399	0.1010	0.0156	22	-0.49	137.05	0
137	Tyrosine (%)	0227	0.6100	0.0200	0.6399	0.1010	0.0156	22	-0.30	137.99	0
137	Tyrosine (%)	0407	0.6140	0.0120	0.6399	0.1010	0.0156	22	-0.26	137.05	0
137	Tyrosine (%)	2188	0.6180	0.0000	0.6399	0.1010	0.0156	22	-0.22	137.05	0
137	Tyrosine (%)	2144	0.6200	0.0000	0.6399	0.1010	0.0156	22	-0.20	137.99	0
137	Tyrosine (%)	0644	0.6470	0.0120	0.6399	0.1010	0.0156	22	0.07	137.00	0
137	Tyrosine (%)	0662	0.6634	0.0315	0.6399	0.1010	0.0156	22	0.23	137.00	0
137	Tyrosine (%)	2059	0.6850	0.0060	0.6399	0.1010	0.0156	22	0.45	137.00	0
137	Tyrosine (%)	0872	0.6890	0.0180	0.6399	0.1010	0.0156	22	0.49	137.00	0
137	Tyrosine (%)	0148	0.6900	0.0200	0.6399	0.1010	0.0156	22	0.50	137.05	0
137	Tyrosine (%)	0918	0.7005	0.0232	0.6399	0.1010	0.0156	22	0.60	137.00	0
137	Tyrosine (%)	0610	0.7150	0.0100	0.6399	0.1010	0.0156	22	0.74	137.05	0
137	Tyrosine (%)	0682	0.7240	0.0000	0.6399	0.1010	0.0156	22	0.83	137.00	0
137	Tyrosine (%)	2196	0.7440	0.0000	0.6399	0.1010	0.0156	22	1.03	137.05	0
137	Tyrosine (%)	0353	0.8350	0.0100	0.6399	0.1010	0.0156	22	1.93	137.00	0
137	Tyrosine (%)	0226	0.8622	0.0674	0.6399	0.1010	0.0156	22	2.20	137.00	0
137	Tyrosine (%)	0941	0.6505	0.0950	0.6399	0.1010	0.0156	22	0.11	137.00	1
138	Valine (%)	0889	0.8075	0.0050	0.9072	0.0521	0.0163	27	-1.91	138.99	0
138	Valine (%)	0682	0.8430	0.0000	0.9072	0.0521	0.0163	27	-1.23	138.00	0
138	Valine (%)	0868	0.8485	0.0250	0.9072	0.0521	0.0163	27	-1.12	138.00	0
138	Valine (%)	0910	0.8600	0.0200	0.9072	0.0521	0.0163	27	-0.90	138.00	0
138	Valine (%)	2196	0.8600	0.0000	0.9072	0.0521	0.0163	27	-0.90	138.05	0
138	Valine (%)	0859	0.8616	0.0132	0.9072	0.0521	0.0163	27	-0.87	138.00	0
138	Valine (%)	0870	0.8620	0.0250	0.9072	0.0521	0.0163	27	-0.87	138.00	0
138	Valine (%)	2246	0.8650	0.0500	0.9072	0.0521	0.0163	27	-0.81	138.05	0
138	Valine (%)	0941	0.8710	0.0180	0.9072	0.0521	0.0163	27	-0.69	138.00	0
138	Valine (%)	0918	0.8761	0.0182	0.9072	0.0521	0.0163	27	-0.60	138.00	0
138	Valine (%)	2188	0.8870	0.0020	0.9072	0.0521	0.0163	27	-0.39	138.05	0
138	Valine (%)	0872	0.8930	0.0440	0.9072	0.0521	0.0163	27	-0.27	138.00	0
138	Valine (%)	0652	0.8950	0.0100	0.9072	0.0521	0.0163	27	-0.23	138.00	0
138	Valine (%)	0619	0.9035	0.0470	0.9072	0.0521	0.0163	27	-0.07	138.00	0
138	Valine (%)	0878	0.9125	0.0230	0.9072	0.0521	0.0163	27	0.10	138.00	0
138	Valine (%)	0571	0.9185	0.0070	0.9072	0.0521	0.0163	27	0.22	138.00	0
138	Valine (%)	0968	0.9350	0.0100	0.9072	0.0521	0.0163	27	0.53	138.00	0
138	Valine (%)	0098	0.9375	0.0030	0.9072	0.0521	0.0163	27	0.58	138.02	0
138	Valine (%)	0227	0.9400	0.0200	0.9072	0.0521	0.0163	27	0.63	138.99	0
138	Valine (%)	0407	0.9450	0.0260	0.9072	0.0521	0.0163	27	0.73	138.05	0
138	Valine (%)	0644	0.9495	0.0070	0.9072	0.0521	0.0163	27	0.81	138.00	0
138	Valine (%)	0148	0.9500	0.0000	0.9072	0.0521	0.0163	27	0.82	138.05	0
138	Valine (%)	2059	0.9615	0.0070	0.9072	0.0521	0.0163	27	1.04	138.00	0
138	Valine (%)	0353	0.9700	0.0000	0.9072	0.0521	0.0163	27	1.21	138.00	0
138	Valine (%)	2144	0.9700	0.0000	0.9072	0.0521	0.0163	27	1.21	138.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
138	Valine (%)	0662	0.9740	0.0492	0.9072	0.0521	0.0163	27	1.28	138.00	0
138	Valine (%)	0610	0.9750	0.0100	0.9072	0.0521	0.0163	27	1.30	138.05	0
138	Valine (%)	0242	0.8000	0.1400	0.9072	0.0521	0.0163	27	-2.06	138.05	1
138	Valine (%)	0226	0.9196	0.1683	0.9072	0.0521	0.0163	27	0.24	138.00	1
139	Taurine (%)	0407	0.0160	0.0020				3		139.05	0
139	Taurine (%)	0889	0.0200	0.0000				3		139.99	0
139	Taurine (%)	0682	0.1020	0.0000				3		139.00	0
139	Taurine (%)	0098	< 0.01					3		139.02	5
139	Taurine (%)	0227	< 0.01					3		139.99	5
160	Fructose (%)	0619	0.1475	0.0010				3		160.99	0
160	Fructose (%)	0297	0.1475	0.0170				3		160.10	0
160	Fructose (%)	0227	0.1950	0.0100				3		160.99	0
161	Galactose (%)	0297	0.0000	0.0000				0		161.10	4
162	Glucose (%)	0619	0.1120	0.0020				2		162.99	0
162	Glucose (%)	0297	0.1690	0.0160				2		162.10	0
162	Glucose (%)	0227	< 0.15					2		162.99	5
163	Lactose (%)	0297	0.0000	0.0000				0		163.10	4
163	Lactose (%)	0619	0.0000	0.0000				0		163.99	4
163	Lactose (%)	0227	< 0.15					0		163.99	5
164	Maltose (%)	0297	0.0000	0.0000				0		164.10	4
164	Maltose (%)	0227	< 0.15					0		164.99	5
165	Sucrose (%)	0297	2.258	0.3050				3		165.10	0
165	Sucrose (%)	0619	2.805	0.0700				3		165.99	0
165	Sucrose (%)	0227	2.860	0.0200				3		165.99	0
166	Raffinose (%)	0619	0.3485	0.0070				3		166.99	0
166	Raffinose (%)	0297	0.5895	0.0890				3		166.10	0
166	Raffinose (%)	0227	0.6700	0.0200				3		166.99	0
167	Stachyose (%)	0619	1.235	0.0300				3		167.99	0
167	Stachyose (%)	0297	1.272	0.1580				3		167.10	0
167	Stachyose (%)	0227	1.500	0.0200				3		167.99	0
345	Amprolium (ppm)	0036	47.55	0.1000	67.23	17.19	3.411	12	-1.14	345.00	0
345	Amprolium (ppm)	0047	52.25	2.300	67.23	17.19	3.411	12	-0.87	345.02	0
345	Amprolium (ppm)	0511	52.55	0.5400	67.23	17.19	3.411	12	-0.85	345.00	0
345	Amprolium (ppm)	0014	56.90	8.800	67.23	17.19	3.411	12	-0.60	345.02	0
345	Amprolium (ppm)	0032	59.50	1.000	67.23	17.19	3.411	12	-0.45	345.02	0
345	Amprolium (ppm)	0033	60.40	1.000	67.23	17.19	3.411	12	-0.40	345.02	0
345	Amprolium (ppm)	0001	67.85	6.100	67.23	17.19	3.411	12	0.04	345.02	0
345	Amprolium (ppm)	0171	70.95	1.300	67.23	17.19	3.411	12	0.22	345.00	0
345	Amprolium (ppm)	0013	71.85	2.900	67.23	17.19	3.411	12	0.27	345.00	0
345	Amprolium (ppm)	0148	85.10	3.200	67.23	17.19	3.411	12	1.04	345.04	0
345	Amprolium (ppm)	0028	88.86	4.550	67.23	17.19	3.411	12	1.26	345.00	0
345	Amprolium (ppm)	2234	121.7	9.137	67.23	17.19	3.411	12	3.17	345.04	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
386	Tiamulin (ppm)	0668	72.70	0.0000	90.66	13.88	1.820	11	-1.29	386.00	0
386	Tiamulin (ppm)	0148	76.40	1.000	90.66	13.88	1.820	11	-1.03	386.02	0
386	Tiamulin (ppm)	2055	83.00	2.000	90.66	13.88	1.820	11	-0.55	386.02	0
386	Tiamulin (ppm)	0001	84.85	4.100	90.66	13.88	1.820	11	-0.42	386.00	0
386	Tiamulin (ppm)	2314	85.78	0.9395	90.66	13.88	1.820	11	-0.35	386.02	0
386	Tiamulin (ppm)	2305	86.33	0.9980	90.66	13.88	1.820	11	-0.31	386.02	0
386	Tiamulin (ppm)	0619	92.60	0.8000	90.66	13.88	1.820	11	0.14	386.02	0
386	Tiamulin (ppm)	2234	97.59	4.080	90.66	13.88	1.820	11	0.50	386.99	0
386	Tiamulin (ppm)	0910	99.15	2.900	90.66	13.88	1.820	11	0.61	386.00	0
386	Tiamulin (ppm)	2144	107.4	2.200	90.66	13.88	1.820	11	1.21	386.02	0
386	Tiamulin (ppm)	2192	111.5	1.000	90.66	13.88	1.820	11	1.50	386.00	0
400	Water Activity (Units)	2174	0.5125	0.0250	0.5436	0.0130	0.0056	9	-2.39	400.01	0
400	Water Activity (Units)	0843	0.5335	0.0010	0.5436	0.0130	0.0056	9	-0.78	400.99	0
400	Water Activity (Units)	2168	0.5365	0.0030	0.5436	0.0130	0.0056	9	-0.55	400.01	0
400	Water Activity (Units)	0843	0.5438	0.0011	0.5436	0.0130	0.0056	9	0.01	400.01	0
400	Water Activity (Units)	2073	0.5440	0.0020	0.5436	0.0130	0.0056	9	0.03	400.99	0
400	Water Activity (Units)	0407	0.5454	0.0012	0.5436	0.0130	0.0056	9	0.14	400.01	0
400	Water Activity (Units)	2303	0.5475	0.0060	0.5436	0.0130	0.0056	9	0.30	400.01	0
400	Water Activity (Units)	0589	0.5550	0.0100	0.5436	0.0130	0.0056	9	0.87	400.01	0
400	Water Activity (Units)	2181	0.5629	0.0012	0.5436	0.0130	0.0056	9	1.48	400.01	0
400	Water Activity (Units)	2109	0.6218	0.0625	0.5436	0.0130	0.0056	9	6.00	400.01	1
516	Arsenic, Total (ppm)	0045	0.0355	0.0070	0.0409	0.0032	0.0092	5		516.00	0
516	Arsenic, Total (ppm)	0096	0.0411	0.0057	0.0409	0.0032	0.0092	5		516.52	0
516	Arsenic, Total (ppm)	0227	0.0415	0.0010	0.0409	0.0032	0.0092	5		516.53	0
516	Arsenic, Total (ppm)	0171	0.0425	0.0290	0.0409	0.0032	0.0092	5		516.00	0
516	Arsenic, Total (ppm)	0553	0.0439	0.0033	0.0409	0.0032	0.0092	5		516.53	0
516	Arsenic, Total (ppm)	0407	1.460	0.1870	0.0409	0.0032	0.0092	5		516.43	1
516	Arsenic, Total (ppm)	2302	< 0.01		0.0409	0.0032	0.0092	5		516.99	5
516	Arsenic, Total (ppm)	0910	< 0.05		0.0409	0.0032	0.0092	5		516.52	5
516	Arsenic, Total (ppm)	0021	< 0.05		0.0409	0.0032	0.0092	5		516.53	5
516	Arsenic, Total (ppm)	0918	< 0.089		0.0409	0.0032	0.0092	5		516.53	5
516	Arsenic, Total (ppm)	0186	< 0.1		0.0409	0.0032	0.0092	5		516.52	5
516	Arsenic, Total (ppm)	0021	< 5		0.0409	0.0032	0.0092	5		516.43	5
516	Arsenic, Total (ppm)	0682	< 10		0.0409	0.0032	0.0092	5		516.43	5
518	Cadmium (ppm)	2302	0.0200	0.0000	0.0630	0.0107	0.0019	11	-4.00	518.99	0
518	Cadmium (ppm)	0407	0.0459	0.0035	0.0630	0.0107	0.0019	11	-1.59	518.43	0
518	Cadmium (ppm)	0171	0.0585	0.0010	0.0630	0.0107	0.0019	11	-0.42	518.41	0
518	Cadmium (ppm)	0227	0.0605	0.0030	0.0630	0.0107	0.0019	11	-0.23	518.53	0
518	Cadmium (ppm)	0021	0.0620	0.0000	0.0630	0.0107	0.0019	11	-0.09	518.53	0
518	Cadmium (ppm)	0096	0.0655	0.0048	0.0630	0.0107	0.0019	11	0.23	518.52	0
518	Cadmium (ppm)	0918	0.0675	0.0003	0.0630	0.0107	0.0019	11	0.42	518.53	0
518	Cadmium (ppm)	0186	0.0675	0.0010	0.0630	0.0107	0.0019	11	0.42	518.52	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
518	Cadmium (ppm)	0407	0.0688	0.0045	0.0630	0.0107	0.0019	11	0.54	518.41	0
518	Cadmium (ppm)	0553	0.0724	0.0020	0.0630	0.0107	0.0019	11	0.88	518.53	0
518	Cadmium (ppm)	0968	0.0765	0.0010	0.0630	0.0107	0.0019	11	1.26	518.43	0
518	Cadmium (ppm)	0910	0.0650	0.0100	0.0630	0.0107	0.0019	11	0.19	518.52	1
518	Cadmium (ppm)	0682	< 0.2		0.0630	0.0107	0.0019	11		518.43	5
518	Cadmium (ppm)	0021	< 2		0.0630	0.0107	0.0019	11		518.43	5
520	Chromium (ppm)	0918	1.383	0.1207	4.566	1.220	0.0842	14	-2.61	520.53	0
520	Chromium (ppm)	0171	3.140	0.0200	4.566	1.220	0.0842	14	-1.17	520.41	0
520	Chromium (ppm)	0186	3.210	0.1000	4.566	1.220	0.0842	14	-1.11	520.52	0
520	Chromium (ppm)	0968	3.750	0.0580	4.566	1.220	0.0842	14	-0.67	520.43	0
520	Chromium (ppm)	0407	3.870	0.0291	4.566	1.220	0.0842	14	-0.57	520.41	0
520	Chromium (ppm)	0096	4.497	0.0731	4.566	1.220	0.0842	14	-0.06	520.52	0
520	Chromium (ppm)	0407	4.665	0.0159	4.566	1.220	0.0842	14	0.08	520.43	0
520	Chromium (ppm)	0693	5.045	0.1030	4.566	1.220	0.0842	14	0.39	520.42	0
520	Chromium (ppm)	0164	5.100	0.1000	4.566	1.220	0.0842	14	0.44	520.53	0
520	Chromium (ppm)	0096	5.127	0.0790	4.566	1.220	0.0842	14	0.46	520.42	0
520	Chromium (ppm)	0553	5.270	0.1400	4.566	1.220	0.0842	14	0.58	520.53	0
520	Chromium (ppm)	0510	5.480	0.0000	4.566	1.220	0.0842	14	0.75	520.43	0
520	Chromium (ppm)	0045	5.640	0.3400	4.566	1.220	0.0842	14	0.88	520.42	0
520	Chromium (ppm)	0682	6.400	0.0000	4.566	1.220	0.0842	14	1.50	520.43	0
520	Chromium (ppm)	0021	6.250	0.9000	4.566	1.220	0.0842	14	1.38	520.53	1
520	Chromium (ppm)	0021	< 22		4.566	1.220	0.0842	14		520.43	5
526	Lead (ppm)	0171	0.0285	0.0110	0.1246	0.0516	0.0193	9	-1.86	526.41	0
526	Lead (ppm)	0096	0.0957	0.0011	0.1246	0.0516	0.0193	9	-0.56	526.52	0
526	Lead (ppm)	0186	0.1050	0.0100	0.1246	0.0516	0.0193	9	-0.38	526.52	0
526	Lead (ppm)	0553	0.1060	0.0100	0.1246	0.0516	0.0193	9	-0.36	526.53	0
526	Lead (ppm)	0910	0.1200	0.0200	0.1246	0.0516	0.0193	9	-0.09	526.52	0
526	Lead (ppm)	0918	0.1248	0.0162	0.1246	0.0516	0.0193	9	0.00	526.53	0
526	Lead (ppm)	0227	0.1435	0.0470	0.1246	0.0516	0.0193	9	0.37	526.53	0
526	Lead (ppm)	0407	0.1774	0.0288	0.1246	0.0516	0.0193	9	1.02	526.41	0
526	Lead (ppm)	0407	0.2887	0.0293	0.1246	0.0516	0.0193	9	3.18	526.43	0
526	Lead (ppm)	2302	< 0.01		0.1246	0.0516	0.0193	9		526.99	5
526	Lead (ppm)	0021	< 0.2		0.1246	0.0516	0.0193	9		526.53	5
526	Lead (ppm)	0682	< 3		0.1246	0.0516	0.0193	9		526.43	5
526	Lead (ppm)	0021	< 6		0.1246	0.0516	0.0193	9		526.43	5
529	Mercury (ppb)	0407	14.85	1.732				1		529.99	0
529	Mercury (ppb)	2302	< 0.01					1		529.99	5
529	Mercury (ppb)	0227	< 10					1		529.99	5
529	Mercury (ppb)	0021	< 44					1		529.99	5
529	Mercury (ppb)	0682	< 10000					1		529.99	5
539	Nickel (ppm)	0186	2.250	0.0400	2.862	0.6367	0.0684	9	-0.96	539.52	0
539	Nickel (ppm)	0407	2.278	0.2054	2.862	0.6367	0.0684	9	-0.92	539.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
539	Nickel (ppm)	0171	2.330	0.0200	2.862	0.6367	0.0684	9	-0.84	539.41	0
539	Nickel (ppm)	0407	2.441	0.0563	2.862	0.6367	0.0684	9	-0.66	539.43	0
539	Nickel (ppm)	0096	2.927	0.0108	2.862	0.6367	0.0684	9	0.10	539.52	0
539	Nickel (ppm)	0918	3.194	0.0228	2.862	0.6367	0.0684	9	0.52	539.53	0
539	Nickel (ppm)	0553	3.220	0.2600	2.862	0.6367	0.0684	9	0.56	539.53	0
539	Nickel (ppm)	0021	3.300	0.0000	2.862	0.6367	0.0684	9	0.69	539.43	0
539	Nickel (ppm)	0682	3.820	0.0000	2.862	0.6367	0.0684	9	1.50	539.43	0
539	Nickel (ppm)	0021	3.400	0.6000	2.862	0.6367	0.0684	9	0.85	539.53	1
702	Butyric Acid (4:0) (%)	0164	< 0.02					0		702.00	5
704	Caproic Acid (6:0) (%)	0164	< 0.02					0		704.00	5
706	Caprylic acid (8:0) (%) (w/w)	0164	< 0.02					0		706.99	5
708	Capric acid (10:0) (%) (w/w)	0164	< 0.02					0		708.99	5
710	Lauric Acid (12:0) (%) (w/w)	0226	< 0.005					0		710.99	5
710	Lauric Acid (12:0) (%) (w/w)	0164	< 0.02					0		710.99	5
714	Myristic Acid (14:0) (%) (w/w)	0226	< 0.005					0		714.99	5
714	Myristic Acid (14:0) (%) (w/w)	0164	< 0.02					0		714.99	5
716	Palmitic Acid (16:0) (%) (w/w)	0226	0.4460	0.0080				2		716.99	0
716	Palmitic Acid (16:0) (%) (w/w)	0164	0.5640	0.0660				2		716.99	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0226	0.0055	0.0010				1		718.99	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0164	< 0.02					1		718.99	5
720	Margaric acid (17:0) (%) (w/w)	0164	< 0.02					0		720.99	5
722	Stearic Acid (18:0) (%) (w/w)	0226	0.0690	0.0020				2		722.99	0
722	Stearic Acid (18:0) (%) (w/w)	0164	0.0830	0.0080				2		722.99	0
724	Oleic Acid (9c-18:1) (%) (w/w)	0226	0.6240	0.0060				2		724.99	0
724	Oleic Acid (9c-18:1) (%) (w/w)	0164	0.6830	0.0560				2		724.99	0
726	Linoleic Acid (9c,12c-18:2) (%) (w/w)	0226	1.484	0.0060				2		726.99	0
726	Linoleic Acid (9c,12c-18:2) (%) (w/w)	0164	1.598	0.1230				2		726.99	0
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (%) (w/v)	0226	0.0915	0.0010				2		728.99	0
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (%) (w/v)	0164	0.1065	0.0110				2		728.99	0
730	Arachidic Acid (20:0) (%) (w/w)	0226	0.0100	0.0000				1		730.99	0
730	Arachidic Acid (20:0) (%) (w/w)	0164	< 0.02					1		730.99	5
732	Gondoic Acid (11c-20:1) (%) (w/w)	0226	0.0100	0.0000				1		732.99	0
732	Gondoic Acid (11c-20:1) (%) (w/w)	0164	< 0.02					1		732.99	5
736	Arachidonic Acid (5c,8c,11c,14c-20:4) (%) (w/v)	0164	< 0.02					0		736.99	5
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c) (%) (w/w)	0226	0.0000	0.0000				0		740.99	4
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c) (%) (w/w)	0164	< 0.02					0		740.99	5
742	Behenic Acid (22:0) (%) (w/w)	0226	0.0070	0.0000				1		742.99	0
742	Behenic Acid (22:0) (%) (w/w)	0164	< 0.02					1		742.99	5
744	Erucic Acid (13c-22:1) (%) (w/w)	0226	< 0.005					0		744.99	5
744	Erucic Acid (13c-22:1) (%) (w/w)	0164	< 0.02					0		744.99	5
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c) (%) (w/w)	0226	0.0000	0.0000				0		746.99	4
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c) (%) (w/w)	0164	< 0.02					0		746.99	5

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
748	Lignoceric Acid (24:0) (% (w/w))	0226	0.0080	0.0000				1	748.99	0	
748	Lignoceric Acid (24:0) (% (w/w))	0164	< 0.02					1	748.99	5	
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16	0226	0.0000	0.0000				0	750.99	4	
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16	0164	< 0.02					0	750.99	5	
752	Nervonic Acid (24:1) isomers (% (w/w))	0226	< 0.005					0	752.99	5	
752	Nervonic Acid (24:1) isomers (% (w/w))	0164	< 0.02					0	752.99	5	
754	Total n-3 Polyunsaturated (Omega-3) Fatty Aci	0164	0.1165	0.0250				1	754.99	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Aci	0164	1.602	0.1260				1	756.99	0	
758	Total Saturated Fatty Acids (% (w/w))	0164	0.7000	0.0820				1	758.99	0	
764	Total cis Monounsaturated Fatty Acids (% (w/w)	0164	0.7355	0.0530				1	764.99	0	
768	Total cis Polyunsaturated Fatty Acids (% (w/w)	0164	1.721	0.1530				1	768.99	0	
770	Total Fat (equivalent to NLEA) (% (w/w))	0164	3.301	0.3000				1	770.99	0	
772	Total Fatty Acids (% (w/w))	0226	2.760	0.0200				2	772.99	0	
772	Total Fatty Acids (% (w/w))	0164	3.156	0.2880				2	772.99	0	

Notes: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2 and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = Rejected for duplicates too far apart, 2 = Rejected as outlier, 5 = A reporting limit and 4 = zeros submitted as values. Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). Flag 3 indicates not used in statistics.