



AAFCO
Proficiency Testing Program



Mycotoxin Contaminants Scheme

Poultry Feed, Medicated

Test Material Code # 202163

Method Proficiency Testing Report

Labs Reporting: 62

Methods Reported 87

Issue Date : 08/31/2021

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	2113	13.50	1.000	16.49	5.628	2.015	11	-0.53	9%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0297	13.80	3.400	16.49	5.628	2.015	11	-0.48	8%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0876	14.05	1.100	16.49	5.628	2.015	11	-0.43	7%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0208	16.15	0.3000	16.49	5.628	2.015	11	-0.06	1%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0034	16.45	2.700	16.49	5.628	2.015	11	-0.01	0%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0027	16.49	4.970	16.49	5.628	2.015	11	0.00	0%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0511	16.50	0.6000	16.49	5.628	2.015	11	0.00	0%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0554	16.95	1.300	16.49	5.628	2.015	11	0.08	1%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	2096	17.40	4.600	16.49	5.628	2.015	11	0.16	3%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0529	19.85	0.1000	16.49	5.628	2.015	11	0.60	10%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	2098	24.15	2.100	16.49	5.628	2.015	11	1.36	23%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0227	< 5		16.49	5.628	2.015	11			5
600.08	Total Aflatoxin, Romer AgraQuant Total Afl 4-40ppb (ppb)	0723	16.35	1.700				2			0
600.08	Total Aflatoxin, Romer AgraQuant Total Afl 4-40ppb (ppb)	2196	20.80	0.0000				2			0
600.12	Total Aflatoxin, r-Biopharm Ridascreen FAST Afl (ppb)	0202	9.940	0.5200				3			0
600.12	Total Aflatoxin, r-Biopharm Ridascreen FAST Afl (ppb)	0590	21.25	0.9000				3			0
600.12	Total Aflatoxin, r-Biopharm Ridascreen FAST Afl (ppb)	2226	23.25	0.0550				3			0
600.13	Total Aflatoxin, r-Biopharm Ridascreen FAST Afl SC (ppb)	0066	13.27	0.8700				3			0
600.13	Total Aflatoxin, r-Biopharm Ridascreen FAST Afl SC (ppb)	0407	23.45	1.626				3			0
600.13	Total Aflatoxin, r-Biopharm Ridascreen FAST Afl SC (ppb)	0035	24.76	0.9900				3			0
600.14	Total Aflatoxin, Vicam Aflatest (ppb)	2268	8.950	0.5000	12.32	4.239	2.140	7	-0.80	14%	0
600.14	Total Aflatoxin, Vicam Aflatest (ppb)	2345	10.00	2.000	12.32	4.239	2.140	7	-0.55	9%	0
600.14	Total Aflatoxin, Vicam Aflatest (ppb)	0190	10.45	1.100	12.32	4.239	2.140	7	-0.44	8%	0
600.14	Total Aflatoxin, Vicam Aflatest (ppb)	0968	11.09	0.3800	12.32	4.239	2.140	7	-0.29	5%	0
600.14	Total Aflatoxin, Vicam Aflatest (ppb)	0004	12.50	1.000	12.32	4.239	2.140	7	0.04	1%	0
600.14	Total Aflatoxin, Vicam Aflatest (ppb)	0520	15.50	1.000	12.32	4.239	2.140	7	0.75	13%	0
600.14	Total Aflatoxin, Vicam Aflatest (ppb)	2199	45.50	9.000	12.32	4.239	2.140	7	7.83	135%	0
600.20	Total Aflatoxin, LC (ppb)	0723	11.05	0.3400	17.46	5.950	1.840	8	-1.08	18%	0
600.20	Total Aflatoxin, LC (ppb)	0968	11.09	0.3800	17.46	5.950	1.840	8	-1.07	18%	0
600.20	Total Aflatoxin, LC (ppb)	2196	12.20	0.0000	17.46	5.950	1.840	8	-0.88	15%	0
600.20	Total Aflatoxin, LC (ppb)	0563	17.50	1.000	17.46	5.950	1.840	8	0.01	0%	0
600.20	Total Aflatoxin, LC (ppb)	2192	18.35	0.5000	17.46	5.950	1.840	8	0.15	3%	0
600.20	Total Aflatoxin, LC (ppb)	0051	18.90	1.600	17.46	5.950	1.840	8	0.24	4%	0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
600.20	Total Aflatoxin, LC (ppb)	0674	23.07	0.0000	17.46	5.950	1.840	8	0.94	16%	0
600.20	Total Aflatoxin, LC (ppb)	0042	28.65	10.90	17.46	5.950	1.840	8	1.88	32%	0
600.21	Total Aflatoxin, LC-PCD FI (ppb)	0010	12.40	3.000				2			0
600.21	Total Aflatoxin, LC-PCD FI (ppb)	0035	23.34	3.110				2			0
600.22	Total Aflatoxin, Vicam Aflatest, LC-FI (ppb)	0098	20.20	1.820				1			0
600.24	Total Aflatoxin, LC-MS/MS (ppb)	0619	11.85	2.500	20.50	6.956	2.207	8	-1.24	21%	0
600.24	Total Aflatoxin, LC-MS/MS (ppb)	0001	15.05	1.100	20.50	6.956	2.207	8	-0.78	13%	0
600.24	Total Aflatoxin, LC-MS/MS (ppb)	0353	17.50	1.000	20.50	6.956	2.207	8	-0.43	7%	0
600.24	Total Aflatoxin, LC-MS/MS (ppb)	0682	17.80	0.0000	20.50	6.956	2.207	8	-0.39	7%	0
600.24	Total Aflatoxin, LC-MS/MS (ppb)	2033	20.25	2.700	20.50	6.956	2.207	8	-0.04	1%	0
600.24	Total Aflatoxin, LC-MS/MS (ppb)	2057	24.66	0.3577	20.50	6.956	2.207	8	0.60	10%	0
600.24	Total Aflatoxin, LC-MS/MS (ppb)	0032	27.95	6.300	20.50	6.956	2.207	8	1.07	18%	0
600.24	Total Aflatoxin, LC-MS/MS (ppb)	2129	28.95	3.700	20.50	6.956	2.207	8	1.21	21%	0
600.98	Total Aflatoxin, Other Rapid Test Kit (ppb)	2050	14.45	1.900				3			0
600.98	Total Aflatoxin, Other Rapid Test Kit (ppb)	2269	15.15	0.3000				3			0
600.98	Total Aflatoxin, Other Rapid Test Kit (ppb)	0142	21.75	4.900				3			0
600.99	Total Aflatoxin, Miscellaneous (ppb)	2302	28.97	0.2290				1			0
601.20	Aflatoxin B1, LC (ppb)	0723	9.990	0.0600	12.56	4.318	0.4280	5			0
601.20	Aflatoxin B1, LC (ppb)	0968	10.21	0.2800	12.56	4.318	0.4280	5			0
601.20	Aflatoxin B1, LC (ppb)	2196	11.39	0.0000	12.56	4.318	0.4280	5			0
601.20	Aflatoxin B1, LC (ppb)	2192	13.70	0.8000	12.56	4.318	0.4280	5			0
601.20	Aflatoxin B1, LC (ppb)	0563	17.50	1.000	12.56	4.318	0.4280	5			0
601.21	Aflatoxin B1, LC-PCD FI (ppb)	0010	11.55	2.900				3			0
601.21	Aflatoxin B1, LC-PCD FI (ppb)	0910	16.69	0.4500				3			0
601.21	Aflatoxin B1, LC-PCD FI (ppb)	0035	21.78	2.990				3			0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0918	10.55	0.1000	17.09	5.827	2.367	19	-1.12	19%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0553	11.07	1.660	17.09	5.827	2.367	19	-1.03	18%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0028	11.74	5.650	17.09	5.827	2.367	19	-0.92	16%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0619	11.85	2.500	17.09	5.827	2.367	19	-0.90	15%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0870	13.73	0.9000	17.09	5.827	2.367	19	-0.58	10%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	2327	14.02	0.0150	17.09	5.827	2.367	19	-0.53	9%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0913	15.01	0.0700	17.09	5.827	2.367	19	-0.36	6%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0964	15.94	0.0600	17.09	5.827	2.367	19	-0.20	3%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0353	16.00	0.0000	17.09	5.827	2.367	19	-0.19	3%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	2053	16.70	0.8000	17.09	5.827	2.367	19	-0.07	1%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0160	17.20	3.200	17.09	5.827	2.367	19	0.02	0%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0682	17.50	0.0000	17.09	5.827	2.367	19	0.07	1%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	2136	18.50	9.000	17.09	5.827	2.367	19	0.24	4%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	2033	20.25	2.700	17.09	5.827	2.367	19	0.54	9%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0043	21.40	1.000	17.09	5.827	2.367	19	0.74	13%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	2140	22.00	4.000	17.09	5.827	2.367	19	0.84	14%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	2057	22.11	0.0199	17.09	5.827	2.367	19	0.86	15%	0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
601.23	Aflatoxin B1, LC-MS/MS (ppb)	2207	24.50	7.000	17.09	5.827	2.367	19	1.27	22%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0032	27.95	6.300	17.09	5.827	2.367	19	1.86	32%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0009	< 3.21		17.09	5.827	2.367	19			5
601.24	Aflatoxin B1, Vicam Aflatest, LC-FI (ppb)	0968	10.21	0.2800				2			0
601.24	Aflatoxin B1, Vicam Aflatest, LC-FI (ppb)	0098	18.62	1.500				2			0
601.98	Aflatoxin B1, Other Rapid Test Kit (ppb)	2355	8.950	0.1000				2			0
601.98	Aflatoxin B1, Other Rapid Test Kit (ppb)	2269	20.95	6.100				2			0
601.99	Aflatoxin B1, Miscellaneous (ppb)	2302	20.50	1.000				1			0
602.20	Aflatoxin B2, LC (ppb)	2196	0.8000	0.0000	0.9113	0.3364	0.0425	4			0
602.20	Aflatoxin B2, LC (ppb)	0968	0.8800	0.1000	0.9113	0.3364	0.0425	4			0
602.20	Aflatoxin B2, LC (ppb)	0723	0.9600	0.0400	0.9113	0.3364	0.0425	4			0
602.20	Aflatoxin B2, LC (ppb)	2192	1.005	0.0300	0.9113	0.3364	0.0425	4			0
602.20	Aflatoxin B2, LC (ppb)	0563	0.0000	0.0000	0.9113	0.3364	0.0425	4			4
602.21	Aflatoxin B2, LC-PCD FI (ppb)	0010	0.8500	0.1000				3			0
602.21	Aflatoxin B2, LC-PCD FI (ppb)	0910	1.195	0.0700				3			0
602.21	Aflatoxin B2, LC-PCD FI (ppb)	0035	1.565	0.1100				3			0
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0870	1.260	0.0400	1.476	0.5378	0.2609	11	-0.40	7%	0
602.23	Aflatoxin B2, LC-MS/MS (ppb)	2053	1.280	0.1600	1.476	0.5378	0.2609	11	-0.36	7%	0
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0353	1.305	0.0900	1.476	0.5378	0.2609	11	-0.32	6%	0
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0160	1.375	0.2100	1.476	0.5378	0.2609	11	-0.19	3%	0
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0913	1.380	0.2600	1.476	0.5378	0.2609	11	-0.18	3%	0
602.23	Aflatoxin B2, LC-MS/MS (ppb)	2327	1.449	0.1230	1.476	0.5378	0.2609	11	-0.05	1%	0
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0553	1.450	0.7000	1.476	0.5378	0.2609	11	-0.05	1%	0
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0682	1.500	0.0000	1.476	0.5378	0.2609	11	0.04	1%	0
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0964	1.635	0.2500	1.476	0.5378	0.2609	11	0.30	5%	0
602.23	Aflatoxin B2, LC-MS/MS (ppb)	2057	1.798	0.3371	1.476	0.5378	0.2609	11	0.60	11%	0
602.23	Aflatoxin B2, LC-MS/MS (ppb)	2136	1.850	0.7000	1.476	0.5378	0.2609	11	0.70	13%	0
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0032	0.0000	0.0000	1.476	0.5378	0.2609	11			4
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0619	0.0000	0.0000	1.476	0.5378	0.2609	11			4
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0918	0.0000	0.0000	1.476	0.5378	0.2609	11			4
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0028	1.070	2.140	1.476	0.5378	0.2609	11			4
602.23	Aflatoxin B2, LC-MS/MS (ppb)	2207	< 1		1.476	0.5378	0.2609	11			5
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0009	< 3.28		1.476	0.5378	0.2609	11			5
602.23	Aflatoxin B2, LC-MS/MS (ppb)	2033	< 5		1.476	0.5378	0.2609	11			5
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0043	< 5.3		1.476	0.5378	0.2609	11			5
602.23	Aflatoxin B2, LC-MS/MS (ppb)	2140	< 20		1.476	0.5378	0.2609	11			5
602.24	Aflatoxin B2, Vicam Aflatest, LC-FI (ppb)	0968	0.8800	0.1000				2			0
602.24	Aflatoxin B2, Vicam Aflatest, LC-FI (ppb)	0098	1.580	0.3200				2			0
603.21	Aflatoxin G1, LC-PCD FI (ppb)	0910	0.2800	0.1000				1			0
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0353	0.2780	0.0000				2			0
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0009	16.70	0.4130				2			0
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0032	0.0000	0.0000				2			4

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0619	0.0000	0.0000				2			4
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0918	0.0000	0.0000				2			4
603.23	Aflatoxin G1, LC-MS/MS (ppb)	2057	< 0.03					2			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	2053	< 0.32					2			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0682	< 0.5					2			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0870	< 0.6					2			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	2136	< 0.83					2			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0160	< 1					2			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0553	< 1					2			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0913	< 1					2			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0964	< 1					2			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	2207	< 1					2			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	2327	< 1.1					2			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	2033	< 5					2			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0043	< 18.2					2			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	2140	< 20					2			5
603.24	Aflatoxin G1, Vicam Aflatest, LC-FI (ppb)	0098	< 15					0			5
603.99	Aflatoxin G1, Miscellaneous (ppb)	2302	3.500	1.000				1			0
604.21	Aflatoxin G2, LC-PCD FI (ppb)	0910	0.0050	0.0000				1			0
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0032	0.0000	0.0000				0			4
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0619	0.0000	0.0000				0			4
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0918	0.0000	0.0000				0			4
604.23	Aflatoxin G2, LC-MS/MS (ppb)	2057	< 0.03					0			5
604.23	Aflatoxin G2, LC-MS/MS (ppb)	2053	< 0.49					0			5
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0682	< 0.5					0			5
604.23	Aflatoxin G2, LC-MS/MS (ppb)	2136	< 0.83					0			5
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0160	< 1					0			5
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0553	< 1					0			5
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0870	< 1					0			5
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0913	< 1					0			5
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0964	< 1					0			5
604.23	Aflatoxin G2, LC-MS/MS (ppb)	2207	< 1					0			5
604.23	Aflatoxin G2, LC-MS/MS (ppb)	2327	< 1.6					0			5
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0009	< 3.26					0			5
604.23	Aflatoxin G2, LC-MS/MS (ppb)	2033	< 5					0			5
604.23	Aflatoxin G2, LC-MS/MS (ppb)	2140	< 20					0			5
604.24	Aflatoxin G2, Vicam Aflatest, LC-FI (ppb)	0098	< 5					0			5
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	2098	1,250	140.0	1,611	485.6	449.0	10	-0.74	11%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0227	1,450	100.0	1,611	485.6	449.0	10	-0.33	5%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0554	1,450	100.0	1,611	485.6	449.0	10	-0.33	5%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0297	1,550	100.0	1,611	485.6	449.0	10	-0.13	2%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0208	1,620	80.00	1,611	485.6	449.0	10	0.02	0%	0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0042	1,650	100.0	1,611	485.6	449.0	10	0.08	1%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0511	1,700	400.0	1,611	485.6	449.0	10	0.18	3%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0529	1,700	400.0	1,611	485.6	449.0	10	0.18	3%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0027	1,800	0.0000	1,611	485.6	449.0	10	0.39	6%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	2113	1,865	3,070	1,611	485.6	449.0	10	0.52	8%	0
610.05	Deoxynivalenol, Romer AgraQuant DON (ppb)	0674	740.0	0.0000				3			0
610.05	Deoxynivalenol, Romer AgraQuant DON (ppb)	0723	1,580	20.00				3			0
610.05	Deoxynivalenol, Romer AgraQuant DON (ppb)	2196	1,850	0.0000				3			0
610.06	Deoxynivalenol, r-Biopharm Ridascreen DON (ppb)	0202	1,290	30.00				1			0
610.07	Deoxynivalenol, r-Biopharm Ridascreen FAST DON (ppb)	0407	1,324	89.00				2			0
610.07	Deoxynivalenol, r-Biopharm Ridascreen FAST DON (ppb)	0590	1,540	20.00				2			0
610.08	Deoxynivalenol, r-Biopharm Ridascreen FAST DON SC (p	0066	1,635	130.0				2			0
610.08	Deoxynivalenol, r-Biopharm Ridascreen FAST DON SC (p	0035	2,112	244.0				2			0
610.20	Deoxynivalenol, LC (ppb)	0098	1,150	100.0	1,270	385.4	50.00	4			0
610.20	Deoxynivalenol, LC (ppb)	0968	1,279	52.00	1,270	385.4	50.00	4			0
610.20	Deoxynivalenol, LC (ppb)	0051	1,303	48.00	1,270	385.4	50.00	4			0
610.20	Deoxynivalenol, LC (ppb)	2196	1,349	0.0000	1,270	385.4	50.00	4			0
610.22	Deoxynivalenol, LC-MS (ppb)	2268	1,370	60.00				1			0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0553	588.4	119.5	1,268	384.7	80.15	19	-1.77	27%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0032	886.0	10.00	1,268	384.7	80.15	19	-0.99	15%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0001	905.0	216.0	1,268	384.7	80.15	19	-0.94	14%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0682	1,040	0.0000	1,268	384.7	80.15	19	-0.59	9%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	2053	1,051	4.000	1,268	384.7	80.15	19	-0.56	9%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	2136	1,100	60.00	1,268	384.7	80.15	19	-0.44	7%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0913	1,170	320.0	1,268	384.7	80.15	19	-0.25	4%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0010	1,215	10.00	1,268	384.7	80.15	19	-0.14	2%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	2033	1,236	54.00	1,268	384.7	80.15	19	-0.08	1%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0353	1,270	54.00	1,268	384.7	80.15	19	0.01	0%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	2140	1,304	66.00	1,268	384.7	80.15	19	0.09	1%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0028	1,305	116.7	1,268	384.7	80.15	19	0.10	1%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	2207	1,405	30.00	1,268	384.7	80.15	19	0.36	5%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0870	1,473	50.00	1,268	384.7	80.15	19	0.53	8%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0918	1,490	59.47	1,268	384.7	80.15	19	0.58	9%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0619	1,539	17.00	1,268	384.7	80.15	19	0.70	11%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	2327	1,556	21.95	1,268	384.7	80.15	19	0.75	11%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0964	1,608	138.3	1,268	384.7	80.15	19	0.88	13%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	2129	1,751	176.0	1,268	384.7	80.15	19	1.26	19%	0
610.25	Deoxynivalenol, GC-MS (ppb)	0563	646.0	72.00				1			0
610.98	Deoxynivalenol, Other Rapid Test Kit (ppb)	2269	1,199	102.0	1,430	432.4	85.50	4			0
610.98	Deoxynivalenol, Other Rapid Test Kit (ppb)	0870	1,420	40.00	1,430	432.4	85.50	4			0
610.98	Deoxynivalenol, Other Rapid Test Kit (ppb)	2050	1,500	200.0	1,430	432.4	85.50	4			0
610.98	Deoxynivalenol, Other Rapid Test Kit (ppb)	0142	1,600	0.0000	1,430	432.4	85.50	4			0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
610.98	Deoxynivalenol, Other Rapid Test Kit (ppb)	2355	1.300	0.0000	1,430	432.4	85.50	4			2
610.99	Deoxynivalenol, Miscellaneous (ppb)	2302	1,553	9.654				2			0
610.99	Deoxynivalenol, Miscellaneous (ppb)	0160	1,746	81.00				2			0
620.01	Total Fumonisin, Neogen Veratox Fumonisin (ppb)	0297	4.000	0.0000	2,291	684.0	75.00	4			0
620.01	Total Fumonisin, Neogen Veratox Fumonisin (ppb)	2113	1,750	100.0	2,291	684.0	75.00	4			0
620.01	Total Fumonisin, Neogen Veratox Fumonisin (ppb)	0027	3,260	100.0	2,291	684.0	75.00	4			0
620.01	Total Fumonisin, Neogen Veratox Fumonisin (ppb)	0511	4,150	100.0	2,291	684.0	75.00	4			0
620.01	Total Fumonisin, Neogen Veratox Fumonisin (ppb)	0554	3,400	600.0	2,291	684.0	75.00	4			1
620.01	Total Fumonisin, Neogen Veratox Fumonisin (ppb)	0227	< 1000		2,291	684.0	75.00	4			5
620.02	Total Fumonisin, Neogen Veratox Fumonisin 5/10 (ppb)	2098	2,925	310.0				2			0
620.02	Total Fumonisin, Neogen Veratox Fumonisin 5/10 (ppb)	0529	4,200	200.0				2			0
620.07	Total Fumonisin, Romer AgraQuant Total Fumonisin (ppb)	0674	3,140	0.0000				2			0
620.07	Total Fumonisin, Romer AgraQuant Total Fumonisin (ppb)	0723	4,095	230.0				2			0
620.09	Total Fumonisin, r-Biopharm Ridascreen FAST FUM (ppb)	2226	2,482	79.00	3,714	1,095	92.80	5			0
620.09	Total Fumonisin, r-Biopharm Ridascreen FAST FUM (ppb)	0035	3,474	113.0	3,714	1,095	92.80	5			0
620.09	Total Fumonisin, r-Biopharm Ridascreen FAST FUM (ppb)	0202	3,577	41.00	3,714	1,095	92.80	5			0
620.09	Total Fumonisin, r-Biopharm Ridascreen FAST FUM (ppb)	0407	3,849	211.0	3,714	1,095	92.80	5			0
620.09	Total Fumonisin, r-Biopharm Ridascreen FAST FUM (ppb)	0066	5,190	20.00	3,714	1,095	92.80	5			0
620.11	Total Fumonisin, Vicam Fumonitest (ppb)	0968	1,744	242.0	2,311	689.9	210.5	4			0
620.11	Total Fumonisin, Vicam Fumonitest (ppb)	0098	1,900	200.0	2,311	689.9	210.5	4			0
620.11	Total Fumonisin, Vicam Fumonitest (ppb)	2268	2,500	0.0000	2,311	689.9	210.5	4			0
620.11	Total Fumonisin, Vicam Fumonitest (ppb)	0190	3,100	400.0	2,311	689.9	210.5	4			0
620.20	Total Fumonisin, LC (ppb)	0051	3,046	31.00				2			0
620.20	Total Fumonisin, LC (ppb)	0208	3,950	500.0				2			0
620.23	Total Fumonisin, LC-MS/MS (ppb)	0619	2,687	660.0	3,474	1,026	292.0	6	-0.77	11%	0
620.23	Total Fumonisin, LC-MS/MS (ppb)	0682	3,092	0.0000	3,474	1,026	292.0	6	-0.37	5%	0
620.23	Total Fumonisin, LC-MS/MS (ppb)	0010	3,150	500.0	3,474	1,026	292.0	6	-0.32	5%	0
620.23	Total Fumonisin, LC-MS/MS (ppb)	0353	3,371	195.0	3,474	1,026	292.0	6	-0.10	1%	0
620.23	Total Fumonisin, LC-MS/MS (ppb)	2033	4,148	127.0	3,474	1,026	292.0	6	0.66	10%	0
620.23	Total Fumonisin, LC-MS/MS (ppb)	0032	4,395	270.0	3,474	1,026	292.0	6	0.90	13%	0
620.98	Total Fumonisin, Other Rapid Test Kit (ppb)	2269	3,621	1,156				2			0
620.98	Total Fumonisin, Other Rapid Test Kit (ppb)	0142	4,550	100.0				2			0
620.99	Total Fumonisin, Miscellaneous (ppb)	0520	3,250	500.0				2			0
620.99	Total Fumonisin, Miscellaneous (ppb)	2302	4,670	7.710				2			0
621.20	Fumonsin B1, LC (ppb)	0563	3,585	50.00				1			0
621.21	Fumonsin B1, LC-FI OPA der. (ppb)	0968	1,342	185.0				2			0
621.21	Fumonsin B1, LC-FI OPA der. (ppb)	0098	1,600	200.0				2			0
621.23	Fumonsin B1, LC-MS/MS (ppb)	2053	1,653	74.00	2,482	739.4	251.2	18	-1.12	17%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	2207	2,000	200.0	2,482	739.4	251.2	18	-0.65	10%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	0619	2,049	563.0	2,482	739.4	251.2	18	-0.59	9%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	0964	2,082	52.55	2,482	739.4	251.2	18	-0.54	8%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	2327	2,175	141.2	2,482	739.4	251.2	18	-0.41	6%	0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
621.23	Fumonsin B1, LC-MS/MS (ppb)	0913	2,180	100.0	2,482	739.4	251.2	18	-0.41	6%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	0010	2,250	500.0	2,482	739.4	251.2	18	-0.31	5%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	0553	2,258	279.2	2,482	739.4	251.2	18	-0.30	5%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	0353	2,397	207.0	2,482	739.4	251.2	18	-0.12	2%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	0682	2,430	0.0000	2,482	739.4	251.2	18	-0.07	1%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	0918	2,469	175.5	2,482	739.4	251.2	18	-0.02	0%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	2140	2,513	293.0	2,482	739.4	251.2	18	0.04	1%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	2136	2,685	170.0	2,482	739.4	251.2	18	0.28	4%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	0870	2,870	20.00	2,482	739.4	251.2	18	0.53	8%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	2033	2,965	90.00	2,482	739.4	251.2	18	0.65	10%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	0032	3,175	230.0	2,482	739.4	251.2	18	0.94	14%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	2129	3,210	1,302	2,482	739.4	251.2	18	0.99	15%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	0160	3,798	124.0	2,482	739.4	251.2	18	1.78	27%	0
621.98	Fumonsin B1, Other Rapid Test Kit (ppb)	2355	3.750	0.9000				1			0
622.21	Fumonisin B2, LC-FI OPA der. (ppb)	0098	300.0	0.0000				2			0
622.21	Fumonisin B2, LC-FI OPA der. (ppb)	0968	402.5	57.00				2			0
622.23	Fumonisin B2, LC-MS/MS (ppb)	2207	450.0	100.0	727.4	224.0	67.09	17	-1.24	19%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	2053	615.5	7.000	727.4	224.0	67.09	17	-0.50	8%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	2136	619.0	62.00	727.4	224.0	67.09	17	-0.48	7%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0619	638.5	97.00	727.4	224.0	67.09	17	-0.40	6%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	2327	644.7	57.30	727.4	224.0	67.09	17	-0.37	6%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0010	650.0	100.0	727.4	224.0	67.09	17	-0.35	5%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0682	660.0	0.0000	727.4	224.0	67.09	17	-0.30	5%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0913	670.0	180.0	727.4	224.0	67.09	17	-0.26	4%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0553	706.6	133.4	727.4	224.0	67.09	17	-0.09	1%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	2140	734.5	69.00	727.4	224.0	67.09	17	0.03	0%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0353	740.0	48.00	727.4	224.0	67.09	17	0.06	1%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0964	783.5	16.53	727.4	224.0	67.09	17	0.25	4%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0870	820.0	0.0000	727.4	224.0	67.09	17	0.41	6%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	2033	855.5	23.00	727.4	224.0	67.09	17	0.57	9%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0032	868.5	25.00	727.4	224.0	67.09	17	0.63	10%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0918	904.5	221.3	727.4	224.0	67.09	17	0.79	12%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0160	1,175	1.000	727.4	224.0	67.09	17	2.00	31%	0
623.23	Fumonisin B3, LC-MS/MS (ppb)	0160	159.0	6.000	288.8	91.20	42.14	10	-1.42	22%	0
623.23	Fumonisin B3, LC-MS/MS (ppb)	2053	189.0	6.000	288.8	91.20	42.14	10	-1.09	17%	0
623.23	Fumonisin B3, LC-MS/MS (ppb)	0010	250.0	100.0	288.8	91.20	42.14	10	-0.43	7%	0
623.23	Fumonisin B3, LC-MS/MS (ppb)	2327	260.6	54.88	288.8	91.20	42.14	10	-0.31	5%	0
623.23	Fumonisin B3, LC-MS/MS (ppb)	0353	278.0	28.00	288.8	91.20	42.14	10	-0.12	2%	0
623.23	Fumonisin B3, LC-MS/MS (ppb)	0913	290.0	20.00	288.8	91.20	42.14	10	0.01	0%	0
623.23	Fumonisin B3, LC-MS/MS (ppb)	2033	327.0	14.00	288.8	91.20	42.14	10	0.42	7%	0
623.23	Fumonisin B3, LC-MS/MS (ppb)	0032	349.5	17.00	288.8	91.20	42.14	10	0.67	11%	0
623.23	Fumonisin B3, LC-MS/MS (ppb)	0870	360.0	0.0000	288.8	91.20	42.14	10	0.78	12%	0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
623.23	Fumonisin B3, LC-MS/MS (ppb)	0553	424.7	175.5	288.8	91.20	42.14	10	1.49	24%	0
630.01	Ochratoxin A, Neogen Veratox Ochratoxin (ppb)	0227	104.5	5.000				3			0
630.01	Ochratoxin A, Neogen Veratox Ochratoxin (ppb)	2098	146.7	5.800				3			0
630.01	Ochratoxin A, Neogen Veratox Ochratoxin (ppb)	0529	659.5	17.00				3			0
630.03	Ochratoxin A, Romer AgraQuant Ochratoxin (ppb)	0723	252.5	5.000				1			0
630.05	Ochratoxin A, r-Biopharm Ridascreen FAST Och A (ppb)	0407	83.39	4.190				3			0
630.05	Ochratoxin A, r-Biopharm Ridascreen FAST Och A (ppb)	2226	101.6	15.29				3			0
630.05	Ochratoxin A, r-Biopharm Ridascreen FAST Och A (ppb)	0066	120.3	6.030				3			0
630.06	Ochratoxin A, Vicam OchraTest (ppb)	0968	133.8	1.980				1			0
630.20	Ochratoxin A, LC (ppb)	0968	133.8	1.980	231.0	73.40	4.240	7	-1.32	21%	0
630.20	Ochratoxin A, LC (ppb)	0674	134.5	0.0000	231.0	73.40	4.240	7	-1.32	21%	0
630.20	Ochratoxin A, LC (ppb)	0563	215.0	18.00	231.0	73.40	4.240	7	-0.22	3%	0
630.20	Ochratoxin A, LC (ppb)	0051	223.5	3.000	231.0	73.40	4.240	7	-0.10	2%	0
630.20	Ochratoxin A, LC (ppb)	0723	263.3	0.7000	231.0	73.40	4.240	7	0.44	7%	0
630.20	Ochratoxin A, LC (ppb)	0098	284.0	6.000	231.0	73.40	4.240	7	0.72	11%	0
630.20	Ochratoxin A, LC (ppb)	2196	363.1	0.0000	231.0	73.40	4.240	7	1.80	29%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0553	67.76	2.760	198.2	63.23	18.79	19	-2.06	33%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0870	128.5	9.400	198.2	63.23	18.79	19	-1.10	18%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0682	138.7	0.0000	198.2	63.23	18.79	19	-0.94	15%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0001	155.5	19.00	198.2	63.23	18.79	19	-0.67	11%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0918	172.6	1.900	198.2	63.23	18.79	19	-0.41	6%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	2207	185.5	39.00	198.2	63.23	18.79	19	-0.20	3%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0353	195.5	11.00	198.2	63.23	18.79	19	-0.04	1%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0619	202.0	8.000	198.2	63.23	18.79	19	0.06	1%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0913	203.2	6.090	198.2	63.23	18.79	19	0.08	1%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	2140	204.0	16.00	198.2	63.23	18.79	19	0.09	1%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0010	208.0	4.000	198.2	63.23	18.79	19	0.16	2%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0032	210.0	40.00	198.2	63.23	18.79	19	0.19	3%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	2129	216.5	21.00	198.2	63.23	18.79	19	0.29	5%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	2327	218.0	6.091	198.2	63.23	18.79	19	0.31	5%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	2136	218.5	89.00	198.2	63.23	18.79	19	0.32	5%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	2033	225.0	30.00	198.2	63.23	18.79	19	0.42	7%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	2053	243.5	15.00	198.2	63.23	18.79	19	0.72	11%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0160	245.0	10.00	198.2	63.23	18.79	19	0.74	12%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0964	246.1	28.78	198.2	63.23	18.79	19	0.76	12%	0
630.98	Ochratoxin A, Other Rapid Test Kit (ppb)	2269	155.6	6.400				1			0
630.99	Ochratoxin A, Miscellaneous (ppb)	2302	123.3	4.341				2			0
630.99	Ochratoxin A, Miscellaneous (ppb)	0520	187.5	5.000				2			0
640.03	T-2 Toxin, Romer AgraQuant T-2 (ppb)	0674	45.23	0.0000				2			0
640.03	T-2 Toxin, Romer AgraQuant T-2 (ppb)	0066	111.0	4.070				2			0
640.05	T-2 Toxin, r-Biopharm Ridascreen FAST T-2 (ppb)	0407	61.95	1.366				1			0
640.21	T-2 Toxin, LC-PCD FI (ppb)	0910	41.50	3.000				1			0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
640.23	T-2 Toxin, LC-MS/MS (ppb)	0918	30.46	6.610	56.01	18.49	5.914	14	-1.38	23%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	0682	40.70	0.0000	56.01	18.49	5.914	14	-0.83	14%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	0619	44.20	4.200	56.01	18.49	5.914	14	-0.64	11%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	2053	49.85	2.100	56.01	18.49	5.914	14	-0.33	5%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	0001	52.15	4.300	56.01	18.49	5.914	14	-0.21	3%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	2033	55.00	10.00	56.01	18.49	5.914	14	-0.05	1%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	0913	55.41	3.590	56.01	18.49	5.914	14	-0.03	1%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	0353	56.50	1.000	56.01	18.49	5.914	14	0.03	0%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	2136	57.35	18.70	56.01	18.49	5.914	14	0.07	1%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	0870	60.55	9.500	56.01	18.49	5.914	14	0.25	4%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	2140	62.50	3.000	56.01	18.49	5.914	14	0.35	6%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	2327	62.85	4.810	56.01	18.49	5.914	14	0.37	6%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	2129	92.00	6.000	56.01	18.49	5.914	14	1.95	32%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	0964	115.3	8.990	56.01	18.49	5.914	14	3.21	53%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	0553	< 1		56.01	18.49	5.914	14			5
640.23	T-2 Toxin, LC-MS/MS (ppb)	2207	< 60		56.01	18.49	5.914	14			5
640.23	T-2 Toxin, LC-MS/MS (ppb)	0160	< 100		56.01	18.49	5.914	14			5
640.25	T-2 Toxin, GC-MS (ppb)	0563	33.50	1.000				1			0
640.98	T-2 Toxin, Other Rapid Test Kit (ppb)	2269	70.60	8.800				2			0
640.98	T-2 Toxin, Other Rapid Test Kit (ppb)	0297	187.5	3.200				2			0
640.99	T-2 Toxin, Miscellaneous (ppb)	2302	101.0	6.000				1			0
641.21	HT-2 Toxin, LC post-col photochem der. FI (ppb)	0910	63.50	3.000				1			0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	0619	52.30	1.600	88.55	28.88	9.902	11	-1.26	20%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	2033	62.00	0.6000	88.55	28.88	9.902	11	-0.92	15%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	0913	72.50	11.25	88.55	28.88	9.902	11	-0.56	9%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	0353	74.00	4.000	88.55	28.88	9.902	11	-0.50	8%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	2053	84.55	1.900	88.55	28.88	9.902	11	-0.14	2%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	2207	85.00	10.00	88.55	28.88	9.902	11	-0.12	2%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	0870	93.50	2.440	88.55	28.88	9.902	11	0.17	3%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	2136	103.5	29.00	88.55	28.88	9.902	11	0.52	8%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	2327	105.2	18.42	88.55	28.88	9.902	11	0.58	9%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	0964	114.5	27.71	88.55	28.88	9.902	11	0.90	15%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	2140	127.0	2.000	88.55	28.88	9.902	11	1.33	22%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	0918	0.0000	0.0000	88.55	28.88	9.902	11			4
642.03	T-2 Toxin + HT-2 Toxin, LC post-col photochem der. FI (ppb)	0964	229.8	36.70				1			0
642.04	T-2 Toxin + HT-2 Toxin, LC-MS (ppb)	0353	130.5	5.000				2			0
642.04	T-2 Toxin + HT-2 Toxin, LC-MS (ppb)	0051	155.5	9.000				2			0
642.21	T-2 Toxin + HT-2 Toxin, Neogen Veratox T-2 + HT-2 (ppb)	2098	190.9	13.10				3			0
642.21	T-2 Toxin + HT-2 Toxin, Neogen Veratox T-2 + HT-2 (ppb)	0529	323.0	2.000				3			0
642.21	T-2 Toxin + HT-2 Toxin, Neogen Veratox T-2 + HT-2 (ppb)	0227	327.5	7.000				3			0
642.98	T-2 Toxin + HT-2 Toxin, Other Rapid Test Kit (ppb)	2269	97.45	12.10	152.7	49.06	19.68	4			0
642.98	T-2 Toxin + HT-2 Toxin, Other Rapid Test Kit (ppb)	0142	143.3	18.60	152.7	49.06	19.68	4			0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
642.98	T-2 Toxin + HT-2 Toxin, Other Rapid Test Kit (ppb)	0202	154.0	4.000	152.7	49.06	19.68	4		0	
642.98	T-2 Toxin + HT-2 Toxin, Other Rapid Test Kit (ppb)	0590	216.0	44.00	152.7	49.06	19.68	4		0	
642.99	T-2 Toxin + HT-2 Toxin, Miscellaneous Technique (ppb)	0407	191.0	6.130				2		0	
642.99	T-2 Toxin + HT-2 Toxin, Miscellaneous Technique (ppb)	2302	243.9	3.027				2		0	
650.01	Zearalenone, Neogen Veratox Zearalenone (ppb)	2098	180.8	24.70	358.8	112.6	46.26	5		0	
650.01	Zearalenone, Neogen Veratox Zearalenone (ppb)	2113	310.0	80.00	358.8	112.6	46.26	5		0	
650.01	Zearalenone, Neogen Veratox Zearalenone (ppb)	0529	357.7	52.10	358.8	112.6	46.26	5		0	
650.01	Zearalenone, Neogen Veratox Zearalenone (ppb)	0027	464.5	35.00	358.8	112.6	46.26	5		0	
650.01	Zearalenone, Neogen Veratox Zearalenone (ppb)	0297	481.1	39.50	358.8	112.6	46.26	5		0	
650.04	Zearalenone, Romer AgraQuant ZON (ppb)	2196	237.0	0.0000				3		0	
650.04	Zearalenone, Romer AgraQuant ZON (ppb)	0674	300.0	0.0000				3		0	
650.04	Zearalenone, Romer AgraQuant ZON (ppb)	0723	334.1	23.10				3		0	
650.06	Zearalenone, r-Biopharm Ridascreen FAST ZON (ppb)	0202	211.0	12.00				3		0	
650.06	Zearalenone, r-Biopharm Ridascreen FAST ZON (ppb)	0590	244.0	34.00				3		0	
650.06	Zearalenone, r-Biopharm Ridascreen FAST ZON (ppb)	2226	253.9	32.39				3		0	
650.07	Zearalenone, r-Biopharm Ridascreen FAST ZON SC (ppb)	0066	146.1	11.30				3		0	
650.07	Zearalenone, r-Biopharm Ridascreen FAST ZON SC (ppb)	0407	147.7	44.31				3		0	
650.07	Zearalenone, r-Biopharm Ridascreen FAST ZON SC (ppb)	0035	209.7	16.15				3		0	
650.20	Zearalenone, LC (ppb)	2196	28.46	0.0000	159.7	51.25	5.043	7	-2.56	41%	0
650.20	Zearalenone, LC (ppb)	0098	125.0	10.00	159.7	51.25	5.043	7	-0.68	11%	0
650.20	Zearalenone, LC (ppb)	0968	138.2	1.800	159.7	51.25	5.043	7	-0.42	7%	0
650.20	Zearalenone, LC (ppb)	0051	173.5	5.000	159.7	51.25	5.043	7	0.27	4%	0
650.20	Zearalenone, LC (ppb)	0563	179.5	13.00	159.7	51.25	5.043	7	0.39	6%	0
650.20	Zearalenone, LC (ppb)	2192	193.5	3.000	159.7	51.25	5.043	7	0.66	11%	0
650.20	Zearalenone, LC (ppb)	0723	262.2	2.500	159.7	51.25	5.043	7	2.00	32%	0
650.21	Zearalenone, LC-FI, ISO (ppb)	0520	104.5	11.00				2		0	
650.21	Zearalenone, LC-FI, ISO (ppb)	0968	138.2	1.800				2		0	
650.23	Zearalenone, LC-MS (ppb)	0353	202.0	2.000				1		0	
650.24	Zearalenone, LC-MS/MS (ppb)	0553	26.85	2.730	181.5	58.05	14.63	18	-2.66	43%	0
650.24	Zearalenone, LC-MS/MS (ppb)	2207	115.0	10.00	181.5	58.05	14.63	18	-1.15	18%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0619	128.0	0.0000	181.5	58.05	14.63	18	-0.92	15%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0032	136.0	10.00	181.5	58.05	14.63	18	-0.78	13%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0001	144.5	17.00	181.5	58.05	14.63	18	-0.64	10%	0
650.24	Zearalenone, LC-MS/MS (ppb)	2033	150.0	0.0000	181.5	58.05	14.63	18	-0.54	9%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0913	154.5	13.71	181.5	58.05	14.63	18	-0.47	7%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0870	168.0	0.4000	181.5	58.05	14.63	18	-0.23	4%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0160	169.5	57.00	181.5	58.05	14.63	18	-0.21	3%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0918	180.4	4.800	181.5	58.05	14.63	18	-0.02	0%	0
650.24	Zearalenone, LC-MS/MS (ppb)	2053	199.5	9.000	181.5	58.05	14.63	18	0.31	5%	0
650.24	Zearalenone, LC-MS/MS (ppb)	2140	200.5	1.000	181.5	58.05	14.63	18	0.33	5%	0
650.24	Zearalenone, LC-MS/MS (ppb)	2136	209.0	46.00	181.5	58.05	14.63	18	0.47	8%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0010	220.0	40.00	181.5	58.05	14.63	18	0.66	11%	0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
650.24	Zearalenone, LC-MS/MS (ppb)	0964	227.0	19.42	181.5	58.05	14.63	18	0.78	13%	0
650.24	Zearalenone, LC-MS/MS (ppb)	2327	235.7	7.215	181.5	58.05	14.63	18	0.93	15%	0
650.24	Zearalenone, LC-MS/MS (ppb)	2129	272.5	25.00	181.5	58.05	14.63	18	1.57	25%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0682	537.7	0.0000	181.5	58.05	14.63	18	6.14	98%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0043	< 356		181.5	58.05	14.63	18			5
650.98	Zearalenone, Other Rapid Test Kit (ppb)	2269	258.4	16.50				2			0
650.98	Zearalenone, Other Rapid Test Kit (ppb)	0142	291.7	18.90				2			0
650.99	Zearalenone, Miscellaneous (ppb)	2268	115.0	10.00				2			0
650.99	Zearalenone, Miscellaneous (ppb)	2302	466.1	0.1310				2			0

Notes: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2 and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = Rejected for duplicates too far apart, 2 = Rejected as outlier, 5 = A reporting limit and 4 = zeros submitted as values. Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). Flag 3 indicates not used in statistics.