



Mycotoxin Contaminants Scheme
Equine Feed

Labs Reporting: 58
Methods Reported 96
Issue Date : 11/30/2021

Test Material Code # 202164

Method Proficiency Testing Report

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	2096	5.850	2.700	7.109	2.482	1.158	9	-0.51	9%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0511	6.000	0.8000	7.109	2.482	1.158	9	-0.45	8%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0529	6.350	0.1000	7.109	2.482	1.158	9	-0.31	5%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0208	6.410	0.8800	7.109	2.482	1.158	9	-0.28	5%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0027	6.880	0.2400	7.109	2.482	1.158	9	-0.09	2%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0876	7.000	0.8000	7.109	2.482	1.158	9	-0.04	1%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0034	7.450	0.3000	7.109	2.482	1.158	9	0.14	2%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0297	8.900	0.6000	7.109	2.482	1.158	9	0.72	13%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0227	17.00	4.000	7.109	2.482	1.158	9	3.99	70%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	2113	< 5		7.109	2.482	1.158	9			5
600.08	Total Aflatoxin, Romer AgraQuant Total Afl 4-40ppb (ppb)	0723	11.75	0.9000				2			0
600.08	Total Aflatoxin, Romer AgraQuant Total Afl 4-40ppb (ppb)	2022	12.72	0.5700				2			0
600.10	Total Aflatoxin, TLC (ppb)	0674	16.15	0.0000				1			0
600.12	Total Aflatoxin, r-Biopharm Ridascreen FAST Afl (ppb)	2226	5.423	0.0640				3			0
600.12	Total Aflatoxin, r-Biopharm Ridascreen FAST Afl (ppb)	0590	6.770	0.2600				3			0
600.12	Total Aflatoxin, r-Biopharm Ridascreen FAST Afl (ppb)	0202	7.600	0.2000				3			0
600.13	Total Aflatoxin, r-Biopharm Ridascreen FAST Afl SC (ppb)	0407	7.140	1.360				3			0
600.13	Total Aflatoxin, r-Biopharm Ridascreen FAST Afl SC (ppb)	0066	7.965	0.0700				3			0
600.13	Total Aflatoxin, r-Biopharm Ridascreen FAST Afl SC (ppb)	0035	10.49	0.0700				3			0
600.14	Total Aflatoxin, Vicam Aflatest (ppb)	0190	3.550	0.7000	4.734	1.671	1.512	6	-0.71	13%	0
600.14	Total Aflatoxin, Vicam Aflatest (ppb)	2268	3.900	0.2000	4.734	1.671	1.512	6	-0.50	9%	0
600.14	Total Aflatoxin, Vicam Aflatest (ppb)	0968	4.195	0.0700	4.734	1.671	1.512	6	-0.32	6%	0
600.14	Total Aflatoxin, Vicam Aflatest (ppb)	0004	4.650	0.7000	4.734	1.671	1.512	6	-0.05	1%	0
600.14	Total Aflatoxin, Vicam Aflatest (ppb)	0520	5.400	0.4000	4.734	1.671	1.512	6	0.40	7%	0
600.14	Total Aflatoxin, Vicam Aflatest (ppb)	2199	37.50	7.000	4.734	1.671	1.512	6	19.61	346%	0
600.20	Total Aflatoxin, LC (ppb)	0968	4.195	0.0700	8.602	2.988	1.223	6	-1.47	26%	0
600.20	Total Aflatoxin, LC (ppb)	0010	7.500	3.000	8.602	2.988	1.223	6	-0.37	6%	0
600.20	Total Aflatoxin, LC (ppb)	0629	8.285	0.9100	8.602	2.988	1.223	6	-0.11	2%	0
600.20	Total Aflatoxin, LC (ppb)	2192	8.740	0.2400	8.602	2.988	1.223	6	0.05	1%	0
600.20	Total Aflatoxin, LC (ppb)	0563	10.24	0.6200	8.602	2.988	1.223	6	0.55	10%	0
600.20	Total Aflatoxin, LC (ppb)	0042	12.65	2.500	8.602	2.988	1.223	6	1.35	24%	0
600.21	Total Aflatoxin, LC-PCD FI (ppb)	0027	6.565	2.910				2			0
600.21	Total Aflatoxin, LC-PCD FI (ppb)	0035	8.285	0.0300				2			0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
600.22	Total Aflatoxin, Vicam Aflatest, LC-FI (ppb)	0098	8.010	1.180				1			0
600.24	Total Aflatoxin, LC-MS/MS (ppb)	0682	5.200	0.0000	8.400	2.920	0.9200	5			0
600.24	Total Aflatoxin, LC-MS/MS (ppb)	2033	7.550	1.500	8.400	2.920	0.9200	5			0
600.24	Total Aflatoxin, LC-MS/MS (ppb)	0353	7.850	1.100	8.400	2.920	0.9200	5			0
600.24	Total Aflatoxin, LC-MS/MS (ppb)	0218	10.20	0.8000	8.400	2.920	0.9200	5			0
600.24	Total Aflatoxin, LC-MS/MS (ppb)	2129	11.20	1.200	8.400	2.920	0.9200	5			0
600.98	Total Aflatoxin, Other Rapid Test Kit (ppb)	2050	6.700	0.0000	7.938	2.763	0.3750	4			0
600.98	Total Aflatoxin, Other Rapid Test Kit (ppb)	2269	6.950	0.5000	7.938	2.763	0.3750	4			0
600.98	Total Aflatoxin, Other Rapid Test Kit (ppb)	0142	8.450	0.9000	7.938	2.763	0.3750	4			0
600.98	Total Aflatoxin, Other Rapid Test Kit (ppb)	2169	9.650	0.1000	7.938	2.763	0.3750	4			0
600.99	Total Aflatoxin, Miscellaneous (ppb)	2302	9.500	1.000				1			0
601.20	Aflatoxin B1, LC (ppb)	0968	3.860	0.0400	6.824	2.385	1.016	5			0
601.20	Aflatoxin B1, LC (ppb)	2192	5.555	1.030	6.824	2.385	1.016	5			0
601.20	Aflatoxin B1, LC (ppb)	0010	6.750	2.500	6.824	2.385	1.016	5			0
601.20	Aflatoxin B1, LC (ppb)	0629	7.715	0.8900	6.824	2.385	1.016	5			0
601.20	Aflatoxin B1, LC (ppb)	0563	10.24	0.6200	6.824	2.385	1.016	5			0
601.21	Aflatoxin B1, LC-PCD FI (ppb)	0027	5.940	2.680				3			0
601.21	Aflatoxin B1, LC-PCD FI (ppb)	0035	7.695	0.0500				3			0
601.21	Aflatoxin B1, LC-PCD FI (ppb)	0910	8.365	1.650				3			0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0043	3.790	1.255	7.653	2.667	0.9857	16	-1.45	25%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0870	5.010	0.2200	7.653	2.667	0.9857	16	-0.99	17%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0682	5.200	0.0000	7.653	2.667	0.9857	16	-0.92	16%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0028	5.770	3.860	7.653	2.667	0.9857	16	-0.71	12%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0553	5.930	0.2800	7.653	2.667	0.9857	16	-0.65	11%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0918	6.050	0.5000	7.653	2.667	0.9857	16	-0.60	10%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0353	7.250	0.9000	7.653	2.667	0.9857	16	-0.15	3%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	2033	7.550	1.500	7.653	2.667	0.9857	16	-0.04	1%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0009	7.750	1.102	7.653	2.667	0.9857	16	0.04	1%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0964	8.655	0.0234	7.653	2.667	0.9857	16	0.38	7%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	2053	8.820	0.2600	7.653	2.667	0.9857	16	0.44	8%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	2327	9.465	0.6700	7.653	2.667	0.9857	16	0.68	12%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0913	9.800	0.4000	7.653	2.667	0.9857	16	0.80	14%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0242	10.00	2.000	7.653	2.667	0.9857	16	0.88	15%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0218	10.20	0.8000	7.653	2.667	0.9857	16	0.95	17%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	2140	11.00	2.000	7.653	2.667	0.9857	16	1.25	22%	0
601.24	Aflatoxin B1, Vicam Aflatest, LC-FI (ppb)	0968	3.860	0.0400				2			0
601.24	Aflatoxin B1, Vicam Aflatest, LC-FI (ppb)	0098	8.010	1.180				2			0
601.91	Aflatoxin B1, Romer AgraQuant Afl B1 (ppb)	2022	7.315	1.110				1			0
601.98	Aflatoxin B1, Other Rapid Test Kit (ppb)	2355	1.550	0.1000				2			0
601.98	Aflatoxin B1, Other Rapid Test Kit (ppb)	2269	6.900	0.6000				2			0
601.99	Aflatoxin B1, Miscellaneous (ppb)	2302	8.000	2.000				1			0
602.20	Aflatoxin B2, LC (ppb)	0968	0.3350	0.0300	0.4338	0.1634	0.0525	4			0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
602.20	Aflatoxin B2, LC (ppb)	2192	0.3800	0.0600	0.4338	0.1634	0.0525	4		0	
602.20	Aflatoxin B2, LC (ppb)	0010	0.4500	0.1000	0.4338	0.1634	0.0525	4		0	
602.20	Aflatoxin B2, LC (ppb)	0629	0.5700	0.0200	0.4338	0.1634	0.0525	4		0	
602.20	Aflatoxin B2, LC (ppb)	0563	0.0000	0.0000	0.4338	0.1634	0.0525	4		4	
602.21	Aflatoxin B2, LC-PCD FI (ppb)	0035	0.5900	0.0200				2		0	
602.21	Aflatoxin B2, LC-PCD FI (ppb)	0910	0.6150	0.0900				2		0	
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0553	0.4550	0.1900				3		0	
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0353	0.5500	0.1000				3		0	
602.23	Aflatoxin B2, LC-MS/MS (ppb)	2053	1.390	0.0000				3		0	
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0043	< 0.0279					3		5	
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0028	< 0.295					3		5	
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0682	< 0.5					3		5	
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0870	< 0.6					3		5	
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0009	< 0.65					3		5	
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0218	< 1					3		5	
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0242	< 1					3		5	
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0913	< 1					3		5	
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0918	< 1					3		5	
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0964	< 1					3		5	
602.23	Aflatoxin B2, LC-MS/MS (ppb)	2033	< 1					3		5	
602.23	Aflatoxin B2, LC-MS/MS (ppb)	2327	< 1.2					3		5	
602.23	Aflatoxin B2, LC-MS/MS (ppb)	2140	< 20					3		5	
602.24	Aflatoxin B2, Vicam Aflatest, LC-FI (ppb)	0968	0.3350	0.0300				1		0	
602.24	Aflatoxin B2, Vicam Aflatest, LC-FI (ppb)	0098	< 1					1		5	
603.20	Aflatoxin G1, LC (ppb)	0629	< 0.25					0		5	
603.21	Aflatoxin G1, LC-PCD FI (ppb)	0910	0.1300	0.1400				1		0	
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0353	< 0.1					0		5	
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0553	< 0.1					0		5	
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0043	< 0.1062					0		5	
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0028	< 0.176					0		5	
603.23	Aflatoxin G1, LC-MS/MS (ppb)	2053	< 0.32					0		5	
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0682	< 0.5					0		5	
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0870	< 0.6					0		5	
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0009	< 0.65					0		5	
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0218	< 1					0		5	
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0242	< 1					0		5	
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0913	< 1					0		5	
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0918	< 1					0		5	
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0964	< 1					0		5	
603.23	Aflatoxin G1, LC-MS/MS (ppb)	2033	< 1					0		5	
603.23	Aflatoxin G1, LC-MS/MS (ppb)	2327	< 1.1					0		5	
603.23	Aflatoxin G1, LC-MS/MS (ppb)	2140	< 20					0		5	

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
603.24	Aflatoxin G1, Vicam Aflatest, LC-FI (ppb)	0098	< 15					0		5	
603.99	Aflatoxin G1, Miscellaneous (ppb)	2302	1.500	1.000				1		0	
604.20	Aflatoxin G2, LC (ppb)	0629	< 0.25					0		5	
604.21	Aflatoxin G2, LC-PCD FI (ppb)	0910	< 0.01					0		5	
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0553	< 0.1					0		5	
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0028	< 0.332					0		5	
604.23	Aflatoxin G2, LC-MS/MS (ppb)	2053	< 0.49					0		5	
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0682	< 0.5					0		5	
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0009	< 0.66					0		5	
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0218	< 1					0		5	
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0242	< 1					0		5	
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0870	< 1					0		5	
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0913	< 1					0		5	
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0918	< 1					0		5	
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0964	< 1					0		5	
604.23	Aflatoxin G2, LC-MS/MS (ppb)	2033	< 1					0		5	
604.23	Aflatoxin G2, LC-MS/MS (ppb)	2327	< 1.6					0		5	
604.23	Aflatoxin G2, LC-MS/MS (ppb)	2140	< 20					0		5	
604.24	Aflatoxin G2, Vicam Aflatest, LC-FI (ppb)	0098	< 5					0		5	
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0227	1,000	0.0000	1,441	435.7	66.25	8	-1.01	15%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0529	1,250	100.0	1,441	435.7	66.25	8	-0.44	7%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0042	1,400	200.0	1,441	435.7	66.25	8	-0.09	1%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	2113	1,400	0.0000	1,441	435.7	66.25	8	-0.09	1%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0511	1,450	100.0	1,441	435.7	66.25	8	0.02	0%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0297	1,550	100.0	1,441	435.7	66.25	8	0.25	4%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	2192	1,625	30.00	1,441	435.7	66.25	8	0.42	6%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0027	1,800	0.0000	1,441	435.7	66.25	8	0.82	12%	0
610.03	Deoxynivalenol, Charm ROSA DON Quant (ppb)	0674	1,330	0.0000				1		0	
610.05	Deoxynivalenol, Romer AgraQuant DON (ppb)	0723	1,595	110.0				2		0	
610.05	Deoxynivalenol, Romer AgraQuant DON (ppb)	2022	1,740	20.00				2		0	
610.06	Deoxynivalenol, r-Biopharm Ridascreen DON (ppb)	0202	1.350	0.1000				1		0	
610.07	Deoxynivalenol, r-Biopharm Ridascreen FAST DON (ppb)	0638	903.0	0.0000				3		0	
610.07	Deoxynivalenol, r-Biopharm Ridascreen FAST DON (ppb)	0590	1,718	15.00				3		0	
610.07	Deoxynivalenol, r-Biopharm Ridascreen FAST DON (ppb)	0407	1,839	38.00				3		0	
610.08	Deoxynivalenol, r-Biopharm Ridascreen FAST DON SC (p	0066	1,525	350.0				2		0	
610.08	Deoxynivalenol, r-Biopharm Ridascreen FAST DON SC (p	0035	2,402	128.0				2		0	
610.20	Deoxynivalenol, LC (ppb)	0910	1.455	0.1700				3		0	
610.20	Deoxynivalenol, LC (ppb)	0968	1,169	17.00				3		0	
610.20	Deoxynivalenol, LC (ppb)	0098	1,300	200.0				3		0	
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0553	835.1	256.0	1,267	384.5	96.31	17	-1.12	17%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0913	900.0	0.0000	1,267	384.5	96.31	17	-0.96	14%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	2053	1,097	16.00	1,267	384.5	96.31	17	-0.44	7%	0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0242	1,196	147.0	1,267	384.5	96.31	17	-0.19	3%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	2055	1,202	41.90	1,267	384.5	96.31	17	-0.17	3%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0353	1,220	168.0	1,267	384.5	96.31	17	-0.12	2%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	2140	1,240	52.00	1,267	384.5	96.31	17	-0.07	1%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0010	1,250	100.0	1,267	384.5	96.31	17	-0.04	1%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0218	1,255	250.0	1,267	384.5	96.31	17	-0.03	0%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0028	1,273	159.5	1,267	384.5	96.31	17	0.01	0%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0682	1,300	0.0000	1,267	384.5	96.31	17	0.09	1%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0918	1,303	0.2300	1,267	384.5	96.31	17	0.09	1%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	2033	1,350	100.0	1,267	384.5	96.31	17	0.22	3%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	2327	1,353	74.00	1,267	384.5	96.31	17	0.22	3%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	2129	1,435	70.00	1,267	384.5	96.31	17	0.44	7%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0964	1,516	162.7	1,267	384.5	96.31	17	0.65	10%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0870	1,620	40.00	1,267	384.5	96.31	17	0.92	14%	0
610.25	Deoxynivalenol, GC-MS (ppb)	0563	2,080	264.0				1			0
610.98	Deoxynivalenol, Other Rapid Test Kit (ppb)	2355	1,150	100.0	1,525	460.4	234.0	5			0
610.98	Deoxynivalenol, Other Rapid Test Kit (ppb)	2269	1,375	70.00	1,525	460.4	234.0	5			0
610.98	Deoxynivalenol, Other Rapid Test Kit (ppb)	2050	1,450	500.0	1,525	460.4	234.0	5			0
610.98	Deoxynivalenol, Other Rapid Test Kit (ppb)	2169	1,700	0.0000	1,525	460.4	234.0	5			0
610.98	Deoxynivalenol, Other Rapid Test Kit (ppb)	0142	1,950	500.0	1,525	460.4	234.0	5			0
610.99	Deoxynivalenol, Miscellaneous (ppb)	0870	1,498	49.90				2			0
610.99	Deoxynivalenol, Miscellaneous (ppb)	2302	1,610	209.0				2			0
620.01	Total Fumonisin, Neogen Veratox Fumonisin (ppb)	0227	1,000	0.0000	1,822	547.3	263.3	6	-1.50	23%	0
620.01	Total Fumonisin, Neogen Veratox Fumonisin (ppb)	2192	1,555	90.00	1,822	547.3	263.3	6	-0.49	7%	0
620.01	Total Fumonisin, Neogen Veratox Fumonisin (ppb)	0511	1,700	200.0	1,822	547.3	263.3	6	-0.22	3%	0
620.01	Total Fumonisin, Neogen Veratox Fumonisin (ppb)	0027	1,825	190.0	1,822	547.3	263.3	6	0.01	0%	0
620.01	Total Fumonisin, Neogen Veratox Fumonisin (ppb)	2113	2,100	600.0	1,822	547.3	263.3	6	0.51	8%	0
620.01	Total Fumonisin, Neogen Veratox Fumonisin (ppb)	0297	2,750	500.0	1,822	547.3	263.3	6	1.70	25%	0
620.01	Total Fumonisin, Neogen Veratox Fumonisin (ppb)	0042	< 2000		1,822	547.3	263.3	6			5
620.02	Total Fumonisin, Neogen Veratox Fumonisin 5/10 (ppb)	0529	2,350	300.0				1			0
620.07	Total Fumonisin, Romer AgraQuant Total Fumonisin (ppb)	0723	2,465	210.0				2			0
620.07	Total Fumonisin, Romer AgraQuant Total Fumonisin (ppb)	2022	4,015	290.0				2			0
620.09	Total Fumonisin, r-Biopharm Ridascreen FAST FUM (ppb)	0202	2,200	0.4000	2,272	678.6	236.7	5			0
620.09	Total Fumonisin, r-Biopharm Ridascreen FAST FUM (ppb)	2226	1,602	60.00	2,272	678.6	236.7	5			0
620.09	Total Fumonisin, r-Biopharm Ridascreen FAST FUM (ppb)	0407	2,865	630.0	2,272	678.6	236.7	5			0
620.09	Total Fumonisin, r-Biopharm Ridascreen FAST FUM (ppb)	0066	3,150	300.0	2,272	678.6	236.7	5			0
620.09	Total Fumonisin, r-Biopharm Ridascreen FAST FUM (ppb)	0035	3,742	193.0	2,272	678.6	236.7	5			0
620.11	Total Fumonisin, Vicam Fumonitest (ppb)	0098	650.0	100.0	1,160	352.7	232.3	4			0
620.11	Total Fumonisin, Vicam Fumonitest (ppb)	0968	659.1	289.0	1,160	352.7	232.3	4			0
620.11	Total Fumonisin, Vicam Fumonitest (ppb)	2268	980.0	40.00	1,160	352.7	232.3	4			0
620.11	Total Fumonisin, Vicam Fumonitest (ppb)	0190	2,350	500.0	1,160	352.7	232.3	4			0
620.12	Total Fumonisin, Vicam Fumonitest 200 (ppb)	0520	1,800	0.0000				1			0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
620.20	Total Fumonisin, LC (ppb)	0208	1,900	0.0000				1			0
620.23	Total Fumonisin, LC-MS/MS (ppb)	2129	1,513	35.00	2,003	600.2	193.8	4			0
620.23	Total Fumonisin, LC-MS/MS (ppb)	0353	2,049	40.00	2,003	600.2	193.8	4			0
620.23	Total Fumonisin, LC-MS/MS (ppb)	0010	2,050	500.0	2,003	600.2	193.8	4			0
620.23	Total Fumonisin, LC-MS/MS (ppb)	2033	2,400	200.0	2,003	600.2	193.8	4			0
620.98	Total Fumonisin, Other Rapid Test Kit (ppb)	0142	2,000	200.0				2			0
620.98	Total Fumonisin, Other Rapid Test Kit (ppb)	2269	3,672	795.0				2			0
620.99	Total Fumonisin, Miscellaneous (ppb)	2302	2,150	188.0				1			0
621.20	Fumonsin B1, LC (ppb)	0563	1,400	90.00				1			0
621.21	Fumonsin B1, LC-FI OPA der. (ppb)	0968	525.0	231.0				2			0
621.21	Fumonsin B1, LC-FI OPA der. (ppb)	0098	650.0	100.0				2			0
621.22	Fumonsin B1, LC-MS (ppb)	0596	2,988	78.41				1			0
621.23	Fumonsin B1, LC-MS/MS (ppb)	0913	1,000	0.0000	1,398	423.1	64.74	16	-0.94	14%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	2053	1,034	28.00	1,398	423.1	64.74	16	-0.86	13%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	0043	1,159	170.0	1,398	423.1	64.74	16	-0.57	9%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	2129	1,171	110.0	1,398	423.1	64.74	16	-0.54	8%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	0964	1,231	4.525	1,398	423.1	64.74	16	-0.40	6%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	0009	1,289	12.36	1,398	423.1	64.74	16	-0.26	4%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	0010	1,350	300.0	1,398	423.1	64.74	16	-0.11	2%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	2327	1,362	13.00	1,398	423.1	64.74	16	-0.09	1%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	2140	1,412	108.0	1,398	423.1	64.74	16	0.03	0%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	0918	1,417	11.53	1,398	423.1	64.74	16	0.04	1%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	0353	1,465	42.00	1,398	423.1	64.74	16	0.16	2%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	0218	1,490	0.0000	1,398	423.1	64.74	16	0.22	3%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	0870	1,570	60.00	1,398	423.1	64.74	16	0.41	6%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	2033	1,750	100.0	1,398	423.1	64.74	16	0.83	13%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	2055	1,996	75.80	1,398	423.1	64.74	16	1.41	21%	0
621.23	Fumonsin B1, LC-MS/MS (ppb)	0553	3,256	0.6100	1,398	423.1	64.74	16	4.39	66%	0
621.98	Fumonsin B1, Other Rapid Test Kit (ppb)	2355	1,900	200.0				1			0
622.21	Fumonisin B2, LC-FI OPA der. (ppb)	0968	134.1	58.00				1			0
622.21	Fumonisin B2, LC-FI OPA der. (ppb)	0098	< 200					1			5
622.22	Fumonisin B2, LC-MS (ppb)	0596	987.1	24.34				1			0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0913	300.0	0.0000	425.1	132.9	42.32	16	-0.94	15%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0043	306.1	24.60	425.1	132.9	42.32	16	-0.90	14%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	2053	308.5	1.000	425.1	132.9	42.32	16	-0.88	14%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	2129	341.5	145.0	425.1	132.9	42.32	16	-0.63	10%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	2327	357.5	7.000	425.1	132.9	42.32	16	-0.51	8%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	2140	381.5	13.00	425.1	132.9	42.32	16	-0.33	5%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0870	410.0	20.00	425.1	132.9	42.32	16	-0.11	2%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0353	421.5	1.000	425.1	132.9	42.32	16	-0.03	0%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0218	426.0	90.00	425.1	132.9	42.32	16	0.01	0%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0009	427.3	5.228	425.1	132.9	42.32	16	0.02	0%	0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
622.23	Fumonisin B2, LC-MS/MS (ppb)	2033	445.0	10.00	425.1	132.9	42.32	16	0.15	2%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0010	450.0	100.0	425.1	132.9	42.32	16	0.19	3%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0964	499.6	25.48	425.1	132.9	42.32	16	0.56	9%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0918	557.6	19.33	425.1	132.9	42.32	16	1.00	16%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	2055	600.1	34.80	425.1	132.9	42.32	16	1.32	21%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0553	714.4	180.7	425.1	132.9	42.32	16	2.18	34%	0
623.23	Fumonisin B3, LC-MS/MS (ppb)	0913	100.0	0.0000	174.4	55.82	9.703	9	-1.33	21%	0
623.23	Fumonisin B3, LC-MS/MS (ppb)	2327	135.0	6.000	174.4	55.82	9.703	9	-0.70	11%	0
623.23	Fumonisin B3, LC-MS/MS (ppb)	2053	156.0	6.000	174.4	55.82	9.703	9	-0.33	5%	0
623.23	Fumonisin B3, LC-MS/MS (ppb)	0353	170.5	3.000	174.4	55.82	9.703	9	-0.07	1%	0
623.23	Fumonisin B3, LC-MS/MS (ppb)	0009	174.0	2.957	174.4	55.82	9.703	9	-0.01	0%	0
623.23	Fumonisin B3, LC-MS/MS (ppb)	0870	190.0	20.00	174.4	55.82	9.703	9	0.28	4%	0
623.23	Fumonisin B3, LC-MS/MS (ppb)	2033	195.0	10.00	174.4	55.82	9.703	9	0.37	6%	0
623.23	Fumonisin B3, LC-MS/MS (ppb)	0010	200.0	0.0000	174.4	55.82	9.703	9	0.46	7%	0
623.23	Fumonisin B3, LC-MS/MS (ppb)	0553	491.4	39.37	174.4	55.82	9.703	9	5.68	91%	0
630.01	Ochratoxin A, Neogen Veratox Ochratoxin (ppb)	0227	59.50	9.000				2			0
630.01	Ochratoxin A, Neogen Veratox Ochratoxin (ppb)	0529	702.8	40.50				2			0
630.03	Ochratoxin A, Romer AgraQuant Ochratoxin (ppb)	0723	229.0	4.000				2			0
630.03	Ochratoxin A, Romer AgraQuant Ochratoxin (ppb)	2022	287.4	0.0300				2			0
630.04	Ochratoxin A, r-Biopharm Ridascreen Ochratoxin A (ppb)	0407	37.28	3.190				1			0
630.05	Ochratoxin A, r-Biopharm Ridascreen FAST Och A (ppb)	2226	138.8	2.454				2			0
630.05	Ochratoxin A, r-Biopharm Ridascreen FAST Och A (ppb)	0066	168.0	8.320				2			0
630.06	Ochratoxin A, Vicam OchraTest (ppb)	0968	140.9	22.45				2			0
630.06	Ochratoxin A, Vicam OchraTest (ppb)	0520	170.0	0.0000				2			0
630.20	Ochratoxin A, LC (ppb)	2192	131.5	9.000	164.5	52.74	17.41	6	-0.63	10%	0
630.20	Ochratoxin A, LC (ppb)	0968	140.9	22.45	164.5	52.74	17.41	6	-0.45	7%	0
630.20	Ochratoxin A, LC (ppb)	0674	148.0	0.0000	164.5	52.74	17.41	6	-0.31	5%	0
630.20	Ochratoxin A, LC (ppb)	0563	173.0	10.00	164.5	52.74	17.41	6	0.16	3%	0
630.20	Ochratoxin A, LC (ppb)	0629	178.0	12.00	164.5	52.74	17.41	6	0.26	4%	0
630.20	Ochratoxin A, LC (ppb)	0098	215.5	51.00	164.5	52.74	17.41	6	0.97	16%	0
630.21	Ochratoxin A, LC-PCD FI (ppb)	0910	191.1	13.50				1			0
630.22	Ochratoxin A, LC-MS (ppb)	0596	193.4	0.9800				1			0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0553	86.54	43.23	181.0	57.88	17.53	15	-1.63	26%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0870	131.0	4.000	181.0	57.88	17.53	15	-0.86	14%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0682	143.2	0.0000	181.0	57.88	17.53	15	-0.65	10%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	2129	158.5	9.000	181.0	57.88	17.53	15	-0.39	6%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0918	163.7	0.3000	181.0	57.88	17.53	15	-0.30	5%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0353	165.0	4.000	181.0	57.88	17.53	15	-0.28	4%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0218	177.5	15.00	181.0	57.88	17.53	15	-0.06	1%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	2053	189.0	10.00	181.0	57.88	17.53	15	0.14	2%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	2140	190.0	14.00	181.0	57.88	17.53	15	0.16	3%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	2327	196.0	36.00	181.0	57.88	17.53	15	0.26	4%	0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
630.23	Ochratoxin A, LC-MS/MS (ppb)	0010	200.0	60.00	181.0	57.88	17.53	15	0.33	5%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0913	203.7	21.20	181.0	57.88	17.53	15	0.39	6%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	2033	215.0	10.00	181.0	57.88	17.53	15	0.59	9%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0964	223.7	18.18	181.0	57.88	17.53	15	0.74	12%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	2055	232.8	18.00	181.0	57.88	17.53	15	0.90	14%	0
630.98	Ochratoxin A, Other Rapid Test Kit (ppb)	2269	172.6	74.80				1			0
630.99	Ochratoxin A, Miscellaneous (ppb)	2302	175.0	20.00				1			0
640.03	T-2 Toxin, Romer AgraQuant T-2 (ppb)	0066	328.8	6.930				1			0
640.05	T-2 Toxin, r-Biopharm Ridascreen FAST T-2 (ppb)	0407	262.9	27.80				1			0
640.21	T-2 Toxin, LC-PCD FI (ppb)	0910	175.0	38.00				1			0
640.23	T-2 Toxin, LC-MS/MS (ppb)	0553	145.7	45.43	314.4	99.06	27.25	14	-1.70	27%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	0682	262.0	0.0000	314.4	99.06	27.25	14	-0.53	8%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	2140	277.0	12.00	314.4	99.06	27.25	14	-0.38	6%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	0353	294.5	29.00	314.4	99.06	27.25	14	-0.20	3%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	2053	296.5	5.000	314.4	99.06	27.25	14	-0.18	3%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	2055	300.8	19.60	314.4	99.06	27.25	14	-0.14	2%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	0870	304.6	23.02	314.4	99.06	27.25	14	-0.10	2%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	2033	310.0	0.0000	314.4	99.06	27.25	14	-0.04	1%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	0913	311.4	9.500	314.4	99.06	27.25	14	-0.03	0%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	2327	332.0	8.000	314.4	99.06	27.25	14	0.18	3%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	0918	334.4	22.63	314.4	99.06	27.25	14	0.20	3%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	0218	361.5	85.00	314.4	99.06	27.25	14	0.48	7%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	0964	423.1	94.32	314.4	99.06	27.25	14	1.10	17%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	2129	441.0	28.00	314.4	99.06	27.25	14	1.28	20%	0
640.25	T-2 Toxin, GC-MS (ppb)	0563	339.0	98.00				1			0
640.98	T-2 Toxin, Other Rapid Test Kit (ppb)	0297	328.5	1.000				2			0
640.98	T-2 Toxin, Other Rapid Test Kit (ppb)	2269	364.8	50.10				2			0
641.21	HT-2 Toxin, LC post-col photochem der. FI (ppb)	0910	56.00	16.00				1			0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	0964	54.24	18.58	97.37	31.67	11.52	11	-1.36	22%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	0918	56.40	0.1900	97.37	31.67	11.52	11	-1.29	21%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	0353	82.00	4.000	97.37	31.67	11.52	11	-0.49	8%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	2033	83.50	13.00	97.37	31.67	11.52	11	-0.44	7%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	2327	88.75	15.10	97.37	31.67	11.52	11	-0.27	4%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	0870	92.20	1.040	97.37	31.67	11.52	11	-0.16	3%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	0913	105.6	0.6000	97.37	31.67	11.52	11	0.26	4%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	2053	116.0	2.000	97.37	31.67	11.52	11	0.59	10%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	0218	120.5	25.00	97.37	31.67	11.52	11	0.73	12%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	2055	125.8	2.200	97.37	31.67	11.52	11	0.90	15%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	2140	152.5	45.00	97.37	31.67	11.52	11	1.74	28%	0
642.04	T-2 Toxin + HT-2 Toxin, LC-MS (ppb)	0353	374.5	29.00				1			0
642.21	T-2 Toxin + HT-2 Toxin, Neogen Veratox T-2 + HT-2 (ppb)	2022	414.0	9.520	719.6	221.7	23.18	4			0
642.21	T-2 Toxin + HT-2 Toxin, Neogen Veratox T-2 + HT-2 (ppb)	2192	671.0	66.00	719.6	221.7	23.18	4			0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
642.21	T-2 Toxin + HT-2 Toxin, Neogen Veratox T-2 + HT-2 (ppb)	0529	671.4	17.20	719.6	221.7	23.18	4			0
642.21	T-2 Toxin + HT-2 Toxin, Neogen Veratox T-2 + HT-2 (ppb)	0227	1,122	0.0000	719.6	221.7	23.18	4			0
642.98	T-2 Toxin + HT-2 Toxin, Other Rapid Test Kit (ppb)	0142	253.6	49.80				3			0
642.98	T-2 Toxin + HT-2 Toxin, Other Rapid Test Kit (ppb)	0590	276.5	25.00				3			0
642.98	T-2 Toxin + HT-2 Toxin, Other Rapid Test Kit (ppb)	2269	503.4	69.10				3			0
642.99	T-2 Toxin + HT-2 Toxin, Miscellaneous Technique (ppb)	0407	251.5	66.46				2			0
642.99	T-2 Toxin + HT-2 Toxin, Miscellaneous Technique (ppb)	2302	450.0	28.00				2			0
650.01	Zearalenone, Neogen Veratox Zearalenone (ppb)	2113	168.0	66.00	363.5	114.1	56.83	4			0
650.01	Zearalenone, Neogen Veratox Zearalenone (ppb)	0529	355.4	78.60	363.5	114.1	56.83	4			0
650.01	Zearalenone, Neogen Veratox Zearalenone (ppb)	0027	455.1	55.70	363.5	114.1	56.83	4			0
650.01	Zearalenone, Neogen Veratox Zearalenone (ppb)	0297	475.5	27.00	363.5	114.1	56.83	4			0
650.02	Zearalenone, Charm ROSA Zearalenone Quant - MeOH (0674	245.6	0.0000				1			0
650.04	Zearalenone, Romer AgraQuant ZON (ppb)	2022	218.0	1.480				2			0
650.04	Zearalenone, Romer AgraQuant ZON (ppb)	0723	267.0	42.00				2			0
650.06	Zearalenone, r-Biopharm Ridascreen FAST ZON (ppb)	0202	157.0	40.00				3			0
650.06	Zearalenone, r-Biopharm Ridascreen FAST ZON (ppb)	2226	196.9	1.019				3			0
650.06	Zearalenone, r-Biopharm Ridascreen FAST ZON (ppb)	0590	226.0	48.00				3			0
650.07	Zearalenone, r-Biopharm Ridascreen FAST ZON SC (ppb)	0407	200.1	16.93				3			0
650.07	Zearalenone, r-Biopharm Ridascreen FAST ZON SC (ppb)	0066	235.5	46.60				3			0
650.07	Zearalenone, r-Biopharm Ridascreen FAST ZON SC (ppb)	0035	253.0	1.190				3			0
650.20	Zearalenone, LC (ppb)	2192	93.95	2.100				3			0
650.20	Zearalenone, LC (ppb)	0563	115.5	3.000				3			0
650.20	Zearalenone, LC (ppb)	0968	140.3	9.120				3			0
650.20	Zearalenone, LC (ppb)	0098	< 100					3			5
650.21	Zearalenone, LC-FI, ISO (ppb)	0968	140.3	9.120				1			0
650.23	Zearalenone, LC-MS (ppb)	0596	107.8	4.330				2			0
650.23	Zearalenone, LC-MS (ppb)	0353	184.5	1.000				2			0
650.24	Zearalenone, LC-MS/MS (ppb)	0553	54.30	20.06	192.7	61.53	14.24	15	-2.25	36%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0043	104.6	10.54	192.7	61.53	14.24	15	-1.43	23%	0
650.24	Zearalenone, LC-MS/MS (ppb)	2053	155.0	2.000	192.7	61.53	14.24	15	-0.61	10%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0870	162.9	5.055	192.7	61.53	14.24	15	-0.48	8%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0682	173.0	0.0000	192.7	61.53	14.24	15	-0.32	5%	0
650.24	Zearalenone, LC-MS/MS (ppb)	2033	180.0	0.0000	192.7	61.53	14.24	15	-0.21	3%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0913	180.3	9.200	192.7	61.53	14.24	15	-0.20	3%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0218	190.5	21.00	192.7	61.53	14.24	15	-0.04	1%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0918	192.9	7.600	192.7	61.53	14.24	15	0.00	0%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0010	200.0	0.0000	192.7	61.53	14.24	15	0.12	2%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0964	219.6	17.68	192.7	61.53	14.24	15	0.44	7%	0
650.24	Zearalenone, LC-MS/MS (ppb)	2327	226.5	11.00	192.7	61.53	14.24	15	0.55	9%	0
650.24	Zearalenone, LC-MS/MS (ppb)	2140	249.5	7.000	192.7	61.53	14.24	15	0.92	15%	0
650.24	Zearalenone, LC-MS/MS (ppb)	2055	269.9	49.50	192.7	61.53	14.24	15	1.25	20%	0
650.24	Zearalenone, LC-MS/MS (ppb)	2129	376.5	53.00	192.7	61.53	14.24	15	2.99	48%	0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
650.24	Zearalenone, LC-MS/MS (ppb)	0242	< 1		192.7	61.53	14.24	15			5
650.98	Zearalenone, Other Rapid Test Kit (ppb)	0142	187.4	15.10				3			0
650.98	Zearalenone, Other Rapid Test Kit (ppb)	2269	264.5	24.70				3			0
650.98	Zearalenone, Other Rapid Test Kit (ppb)	2169	297.6	2.000				3			0
650.99	Zearalenone, Miscellaneous (ppb)	2268	102.5	15.00				3			0
650.99	Zearalenone, Miscellaneous (ppb)	0520	120.0	0.0000				3			0
650.99	Zearalenone, Miscellaneous (ppb)	2302	190.0	10.00				3			0

Notes: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2 and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = Rejected for duplicates too far apart, 2 = Rejected as outlier, 5 = A reporting limit and 4 = zeros submitted as values. Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). Flag 3 indicates not used in statistics.