



**AAFCO**  
Association of American Feed Control Officials



**Mycotoxin Contaminants Scheme**

**Pea Protein**

**Test Material Code # 202261**

**Method Proficiency Testing Report**

**# Labs Reporting: 57**

**# Methods Reported 80**

**Issue Date : 02/28/2022**

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	2096	4.250	0.7000	6.036	2.117	1.027	7	-0.84	15%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0297	4.600	0.8000	6.036	2.117	1.027	7	-0.68	12%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0511	4.850	0.5000	6.036	2.117	1.027	7	-0.56	10%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	2098	4.990	0.4000	6.036	2.117	1.027	7	-0.49	9%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0027	6.815	2.090	6.036	2.117	1.027	7	0.37	6%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0034	7.600	1.600	6.036	2.117	1.027	7	0.74	13%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0014	9.150	1.100	6.036	2.117	1.027	7	1.47	26%	0
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0227	< 5		6.036	2.117	1.027	7			5
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0425	< 5		6.036	2.117	1.027	7			5
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	0876	< 5		6.036	2.117	1.027	7			5
600.01	Total Aflatoxin, Neogen Veratox Aflatoxin (ppb)	2113	< 5		6.036	2.117	1.027	7			5
600.08	Total Aflatoxin, Romer AgraQuant Total Afl 4-40ppb (ppb)	0723	6.350	0.1000				2			0
600.08	Total Aflatoxin, Romer AgraQuant Total Afl 4-40ppb (ppb)	2196	6.910	0.0000				2			0
600.10	Total Aflatoxin, TLC (ppb)	0674	4.400	0.0000				1			0
600.12	Total Aflatoxin, r-Biopharm Ridascreen FAST Afl (ppb)	2226	2.281	0.2610				3			0
600.12	Total Aflatoxin, r-Biopharm Ridascreen FAST Afl (ppb)	0202	4.980	0.2800				3			0
600.12	Total Aflatoxin, r-Biopharm Ridascreen FAST Afl (ppb)	2066	5.745	0.3100				3			0
600.13	Total Aflatoxin, r-Biopharm Ridascreen FAST Afl SC (ppb)	0066	2.425	0.3500				2			0
600.13	Total Aflatoxin, r-Biopharm Ridascreen FAST Afl SC (ppb)	0407	3.625	1.130				2			0
600.14	Total Aflatoxin, Vicam Aflatest (ppb)	2369	3.495	0.1500				3			0
600.14	Total Aflatoxin, Vicam Aflatest (ppb)	0520	3.600	0.0000				3			0
600.14	Total Aflatoxin, Vicam Aflatest (ppb)	0004	4.600	0.2000				3			0
600.20	Total Aflatoxin, LC (ppb)	0723	2.280	0.0200	4.591	1.622	0.4100	9	-1.42	25%	0
600.20	Total Aflatoxin, LC (ppb)	2196	3.000	0.0000	4.591	1.622	0.4100	9	-0.98	17%	0
600.20	Total Aflatoxin, LC (ppb)	2369	3.495	0.1500	4.591	1.622	0.4100	9	-0.68	12%	0
600.20	Total Aflatoxin, LC (ppb)	0010	4.000	0.0000	4.591	1.622	0.4100	9	-0.36	6%	0
600.20	Total Aflatoxin, LC (ppb)	0563	4.920	0.1400	4.591	1.622	0.4100	9	0.20	4%	0
600.20	Total Aflatoxin, LC (ppb)	0629	5.450	0.9000	4.591	1.622	0.4100	9	0.53	9%	0
600.20	Total Aflatoxin, LC (ppb)	0051	5.800	2.000	4.591	1.622	0.4100	9	0.75	13%	0
600.20	Total Aflatoxin, LC (ppb)	2066	5.825	0.2500	4.591	1.622	0.4100	9	0.76	13%	0
600.20	Total Aflatoxin, LC (ppb)	2192	6.545	0.2300	4.591	1.622	0.4100	9	1.20	21%	0
600.22	Total Aflatoxin, Vicam Aflatest, LC-FI (ppb)	0098	5.900	0.1000				1			0
600.24	Total Aflatoxin, LC-MS/MS (ppb)	2057	4.273	0.2709	6.196	2.171	0.6160	4			0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
600.24	Total Aflatoxin, LC-MS/MS (ppb)	0353	4.361	0.8930	6.196	2.171	0.6160	4			0
600.24	Total Aflatoxin, LC-MS/MS (ppb)	0682	5.700	0.0000	6.196	2.171	0.6160	4			0
600.24	Total Aflatoxin, LC-MS/MS (ppb)	2129	10.45	1.300	6.196	2.171	0.6160	4			0
600.24	Total Aflatoxin, LC-MS/MS (ppb)	0032	< 10		6.196	2.171	0.6160	4			5
600.98	Total Aflatoxin, Other Rapid Test Kit (ppb)	2388	4.300	1.000	5.225	1.840	0.6000	4			0
600.98	Total Aflatoxin, Other Rapid Test Kit (ppb)	2050	4.600	0.2000	5.225	1.840	0.6000	4			0
600.98	Total Aflatoxin, Other Rapid Test Kit (ppb)	2269	5.650	0.5000	5.225	1.840	0.6000	4			0
600.98	Total Aflatoxin, Other Rapid Test Kit (ppb)	0142	6.350	0.7000	5.225	1.840	0.6000	4			0
600.99	Total Aflatoxin, Miscellaneous (ppb)	2302	5.500	1.000				1			0
601.20	Aflatoxin B1, LC (ppb)	0723	2.130	0.0200	4.256	1.507	0.5321	9	-1.41	25%	0
601.20	Aflatoxin B1, LC (ppb)	2196	3.000	0.0000	4.256	1.507	0.5321	9	-0.83	15%	0
601.20	Aflatoxin B1, LC (ppb)	2066	3.055	0.0100	4.256	1.507	0.5321	9	-0.80	14%	0
601.20	Aflatoxin B1, LC (ppb)	2369	3.324	0.1590	4.256	1.507	0.5321	9	-0.62	11%	0
601.20	Aflatoxin B1, LC (ppb)	0010	4.000	0.2000	4.256	1.507	0.5321	9	-0.17	3%	0
601.20	Aflatoxin B1, LC (ppb)	0563	4.920	0.1400	4.256	1.507	0.5321	9	0.44	8%	0
601.20	Aflatoxin B1, LC (ppb)	0629	5.100	0.8000	4.256	1.507	0.5321	9	0.56	10%	0
601.20	Aflatoxin B1, LC (ppb)	2192	5.880	0.3000	4.256	1.507	0.5321	9	1.08	19%	0
601.20	Aflatoxin B1, LC (ppb)	0646	44.95	3.160	4.256	1.507	0.5321	9	27.01	478%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	2057	4.105	0.2550	4.969	1.752	0.8413	15	-0.49	9%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0964	4.107	0.0133	4.969	1.752	0.8413	15	-0.49	9%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0160	4.160	1.400	4.969	1.752	0.8413	15	-0.46	8%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	2136	4.280	0.4000	4.969	1.752	0.8413	15	-0.39	7%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0353	4.361	0.8930	4.969	1.752	0.8413	15	-0.35	6%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	2140	4.500	1.000	4.969	1.752	0.8413	15	-0.27	5%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0870	4.980	0.4400	4.969	1.752	0.8413	15	0.01	0%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	2207	5.000	0.0000	4.969	1.752	0.8413	15	0.02	0%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	2053	5.005	0.3300	4.969	1.752	0.8413	15	0.02	0%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0009	5.024	0.8270	4.969	1.752	0.8413	15	0.03	1%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0913	5.400	0.2000	4.969	1.752	0.8413	15	0.25	4%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	2327	5.530	0.2800	4.969	1.752	0.8413	15	0.32	6%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0682	5.700	0.0000	4.969	1.752	0.8413	15	0.42	7%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0043	6.921	2.951	4.969	1.752	0.8413	15	1.11	20%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0028	7.175	3.630	4.969	1.752	0.8413	15	1.26	22%	0
601.23	Aflatoxin B1, LC-MS/MS (ppb)	0032	< 10		4.969	1.752	0.8413	15			5
601.24	Aflatoxin B1, Vicam Aflatest, LC-FI (ppb)	2369	3.324	0.1590				2			0
601.24	Aflatoxin B1, Vicam Aflatest, LC-FI (ppb)	0098	5.900	0.1000				2			0
601.98	Aflatoxin B1, Other Rapid Test Kit (ppb)	2269	6.750	1.700				1			0
602.20	Aflatoxin B2, LC (ppb)	2066	0.1350	0.0500	0.2563	0.0979	0.0290	6	-1.24	24%	0
602.20	Aflatoxin B2, LC (ppb)	0723	0.1500	0.0000	0.2563	0.0979	0.0290	6	-1.09	21%	0
602.20	Aflatoxin B2, LC (ppb)	2369	0.1730	0.0040	0.2563	0.0979	0.0290	6	-0.85	16%	0
602.20	Aflatoxin B2, LC (ppb)	0010	0.3000	0.0000	0.2563	0.0979	0.0290	6	0.45	9%	0
602.20	Aflatoxin B2, LC (ppb)	0629	0.3500	0.1000	0.2563	0.0979	0.0290	6	0.96	18%	0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
602.20	Aflatoxin B2, LC (ppb)	2192	0.4300	0.0200	0.2563	0.0979	0.0290	6	1.77	34%	0
602.20	Aflatoxin B2, LC (ppb)	0563	< 5		0.2563	0.0979	0.0290	6			5
602.23	Aflatoxin B2, LC-MS/MS (ppb)	2057	0.1677	0.0160				2			0
602.23	Aflatoxin B2, LC-MS/MS (ppb)	2327	0.3830	0.0080				2			0
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0032	0.0000	0.0000				2			4
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0043	< 0.0279					2			5
602.23	Aflatoxin B2, LC-MS/MS (ppb)	2053	< 0.18					2			5
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0028	< 0.295					2			5
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0682	< 0.5					2			5
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0870	< 0.6					2			5
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0009	< 0.65					2			5
602.23	Aflatoxin B2, LC-MS/MS (ppb)	2136	< 0.83					2			5
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0160	< 1					2			5
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0913	< 1					2			5
602.23	Aflatoxin B2, LC-MS/MS (ppb)	0964	< 1					2			5
602.23	Aflatoxin B2, LC-MS/MS (ppb)	2207	< 1					2			5
602.23	Aflatoxin B2, LC-MS/MS (ppb)	2140	< 20					2			5
602.24	Aflatoxin B2, Vicam Aflatest, LC-FI (ppb)	2369	0.1730	0.0040				1			0
602.24	Aflatoxin B2, Vicam Aflatest, LC-FI (ppb)	0098	< 1					1			5
603.20	Aflatoxin G1, LC (ppb)	2066	2.635	0.1900				1			0
603.20	Aflatoxin G1, LC (ppb)	0629	< 0.25					1			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0043	0.3030	0.0877				1			0
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0032	0.0000	0.0000				1			4
603.23	Aflatoxin G1, LC-MS/MS (ppb)	2057	< 0.03					1			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0028	< 0.176					1			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	2053	< 0.32					1			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0682	< 0.5					1			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0870	< 0.6					1			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0009	< 0.65					1			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	2136	< 0.83					1			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0160	< 1					1			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0913	< 1					1			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	0964	< 1					1			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	2207	< 1					1			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	2327	< 1.1					1			5
603.23	Aflatoxin G1, LC-MS/MS (ppb)	2140	< 20					1			5
603.24	Aflatoxin G1, Vicam Aflatest, LC-FI (ppb)	0098	< 15					0			5
604.20	Aflatoxin G2, LC (ppb)	0629	< 0.25					0			5
604.21	Aflatoxin G2, LC-PCD FI (ppb)	0910	< 0.01					0			5
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0032	0.0000	0.0000				0			4
604.23	Aflatoxin G2, LC-MS/MS (ppb)	2057	< 0.03					0			5
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0028	< 0.332					0			5

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
604.23	Aflatoxin G2, LC-MS/MS (ppb)	2053	< 0.49					0		5	
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0682	< 0.5					0		5	
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0009	< 0.66					0		5	
604.23	Aflatoxin G2, LC-MS/MS (ppb)	2136	< 0.83					0		5	
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0160	< 1					0		5	
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0870	< 1					0		5	
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0913	< 1					0		5	
604.23	Aflatoxin G2, LC-MS/MS (ppb)	0964	< 1					0		5	
604.23	Aflatoxin G2, LC-MS/MS (ppb)	2207	< 1					0		5	
604.23	Aflatoxin G2, LC-MS/MS (ppb)	2327	< 1.6					0		5	
604.23	Aflatoxin G2, LC-MS/MS (ppb)	2140	< 20					0		5	
604.24	Aflatoxin G2, Vicam Aflatest, LC-FI (ppb)	0098	< 5					0		5	
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0227	900.0	0.0000	1,243	377.4	116.0	10	-0.91	14%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0425	1,000	0.0000	1,243	377.4	116.0	10	-0.64	10%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0027	1,100	200.0	1,243	377.4	116.0	10	-0.38	6%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0297	1,200	0.0000	1,243	377.4	116.0	10	-0.11	2%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0511	1,200	200.0	1,243	377.4	116.0	10	-0.11	2%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	2098	1,275	70.00	1,243	377.4	116.0	10	0.08	1%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0208	1,370	80.00	1,243	377.4	116.0	10	0.34	5%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	2192	1,435	110.0	1,243	377.4	116.0	10	0.51	8%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0042	1,450	300.0	1,243	377.4	116.0	10	0.55	8%	0
610.01	Deoxynivalenol, Neogen Veratox for DON (ppb)	0014	1,500	200.0	1,243	377.4	116.0	10	0.68	10%	0
610.05	Deoxynivalenol, Romer AgraQuant DON (ppb)	0674	660.0	0.0000				3			0
610.05	Deoxynivalenol, Romer AgraQuant DON (ppb)	0723	1,450	100.0				3			0
610.05	Deoxynivalenol, Romer AgraQuant DON (ppb)	2196	1,558	0.0000				3			0
610.06	Deoxynivalenol, r-Biopharm Ridascreen DON (ppb)	0202	920.0	0.0000				1			0
610.07	Deoxynivalenol, r-Biopharm Ridascreen FAST DON (ppb)	2066	935.0	50.00				3			0
610.07	Deoxynivalenol, r-Biopharm Ridascreen FAST DON (ppb)	0407	1,041	81.00				3			0
610.07	Deoxynivalenol, r-Biopharm Ridascreen FAST DON (ppb)	0590	1,079	362.0				3			0
610.08	Deoxynivalenol, r-Biopharm Ridascreen FAST DON SC (p	0066	1,015	230.0				2			0
610.08	Deoxynivalenol, r-Biopharm Ridascreen FAST DON SC (p	0035	1,496	153.0				2			0
610.20	Deoxynivalenol, LC (ppb)	0098	800.0	0.0000				3			0
610.20	Deoxynivalenol, LC (ppb)	2196	937.0	0.0000				3			0
610.20	Deoxynivalenol, LC (ppb)	0051	1,019	3.000				3			0
610.20	Deoxynivalenol, LC (ppb)	2369	1,137	20.00				3			1
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0964	231.2	75.66	957.8	292.8	63.43	14	-2.48	38%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0043	626.4	50.85	957.8	292.8	63.43	14	-1.13	17%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	2136	717.0	132.0	957.8	292.8	63.43	14	-0.82	13%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0913	850.0	100.0	957.8	292.8	63.43	14	-0.37	6%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	2140	861.0	118.0	957.8	292.8	63.43	14	-0.33	5%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0032	902.5	25.00	957.8	292.8	63.43	14	-0.19	3%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0028	930.8	228.5	957.8	292.8	63.43	14	-0.09	1%	0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
610.23	Deoxynivalenol, LC-MS/MS (ppb)	2207	950.0	20.00	957.8	292.8	63.43	14	-0.03	0%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0353	1,059	64.00	957.8	292.8	63.43	14	0.35	5%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0682	1,110	0.0000	957.8	292.8	63.43	14	0.52	8%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	2053	1,164	19.00	957.8	292.8	63.43	14	0.70	11%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0870	1,170	20.00	957.8	292.8	63.43	14	0.72	11%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	2327	1,235	32.00	957.8	292.8	63.43	14	0.95	14%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	2129	1,245	3.000	957.8	292.8	63.43	14	0.98	15%	0
610.23	Deoxynivalenol, LC-MS/MS (ppb)	0010	< 150		957.8	292.8	63.43	14			5
610.25	Deoxynivalenol, GC-MS (ppb)	0563	1,568	110.0				1			0
610.98	Deoxynivalenol, Other Rapid Test Kit (ppb)	2269	1,140	180.0				3			0
610.98	Deoxynivalenol, Other Rapid Test Kit (ppb)	2050	1,550	100.0				3			0
610.98	Deoxynivalenol, Other Rapid Test Kit (ppb)	0142	1,800	0.0000				3			0
610.99	Deoxynivalenol, Miscellaneous (ppb)	0870	1,136	25.00				3			0
610.99	Deoxynivalenol, Miscellaneous (ppb)	0160	1,200	0.0000				3			0
610.99	Deoxynivalenol, Miscellaneous (ppb)	2302	1,733	10.00				3			0
620.01	Total Fumonisin, Neogen Veratox Fumonisin (ppb)	0511	450.0	100.0	662.5	204.6	135.0	6	-1.04	16%	0
620.01	Total Fumonisin, Neogen Veratox Fumonisin (ppb)	0027	645.0	290.0	662.5	204.6	135.0	6	-0.09	1%	0
620.01	Total Fumonisin, Neogen Veratox Fumonisin (ppb)	0042	650.0	100.0	662.5	204.6	135.0	6	-0.06	1%	0
620.01	Total Fumonisin, Neogen Veratox Fumonisin (ppb)	0297	650.0	100.0	662.5	204.6	135.0	6	-0.06	1%	0
620.01	Total Fumonisin, Neogen Veratox Fumonisin (ppb)	0014	700.0	0.0000	662.5	204.6	135.0	6	0.18	3%	0
620.01	Total Fumonisin, Neogen Veratox Fumonisin (ppb)	2192	880.0	220.0	662.5	204.6	135.0	6	1.06	16%	0
620.01	Total Fumonisin, Neogen Veratox Fumonisin (ppb)	0227	< 1000		662.5	204.6	135.0	6			5
620.01	Total Fumonisin, Neogen Veratox Fumonisin (ppb)	0425	< 1000		662.5	204.6	135.0	6			5
620.01	Total Fumonisin, Neogen Veratox Fumonisin (ppb)	2113	< 1000		662.5	204.6	135.0	6			5
620.02	Total Fumonisin, Neogen Veratox Fumonisin 5/10 (ppb)	2098	780.0	120.0				1			0
620.07	Total Fumonisin, Romer AgraQuant Total Fumonisin (ppb)	0674	720.0	0.0000				3			0
620.07	Total Fumonisin, Romer AgraQuant Total Fumonisin (ppb)	2196	924.3	0.0000				3			0
620.07	Total Fumonisin, Romer AgraQuant Total Fumonisin (ppb)	0723	1,055	30.00				3			0
620.09	Total Fumonisin, r-Biopharm Ridascreen FAST FUM (ppb)	0407	537.5	5.000	815.3	250.3	99.67	6	-1.11	17%	0
620.09	Total Fumonisin, r-Biopharm Ridascreen FAST FUM (ppb)	2226	595.5	103.0	815.3	250.3	99.67	6	-0.88	13%	0
620.09	Total Fumonisin, r-Biopharm Ridascreen FAST FUM (ppb)	0035	735.0	4.000	815.3	250.3	99.67	6	-0.32	5%	0
620.09	Total Fumonisin, r-Biopharm Ridascreen FAST FUM (ppb)	0202	756.0	46.00	815.3	250.3	99.67	6	-0.24	4%	0
620.09	Total Fumonisin, r-Biopharm Ridascreen FAST FUM (ppb)	2066	980.0	120.0	815.3	250.3	99.67	6	0.66	10%	0
620.09	Total Fumonisin, r-Biopharm Ridascreen FAST FUM (ppb)	0066	1,340	320.0	815.3	250.3	99.67	6	2.10	32%	0
620.11	Total Fumonisin, Vicam Fumonitest (ppb)	0098	< 200					0			5
620.12	Total Fumonisin, Vicam Fumonitest 200 (ppb)	0520	0.6850	0.0300				1			0
620.23	Total Fumonisin, LC-MS/MS (ppb)	0010	700.0	200.0	762.8	240.9	105.5	4			0
620.23	Total Fumonisin, LC-MS/MS (ppb)	0682	780.0	0.0000	762.8	240.9	105.5	4			0
620.23	Total Fumonisin, LC-MS/MS (ppb)	0032	784.5	97.00	762.8	240.9	105.5	4			0
620.23	Total Fumonisin, LC-MS/MS (ppb)	0353	786.5	125.0	762.8	240.9	105.5	4			0
620.23	Total Fumonisin, LC-MS/MS (ppb)	2129	1,830	108.0	762.8	240.9	105.5	4			2
620.98	Total Fumonisin, Other Rapid Test Kit (ppb)	2388	850.0	300.0	1,331	403.2	116.0	4			0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
620.98	Total Fumonisin, Other Rapid Test Kit (ppb)	0142	1,200	0.0000	1,331	403.2	116.0	4			0
620.98	Total Fumonisin, Other Rapid Test Kit (ppb)	2387	1,250	100.0	1,331	403.2	116.0	4			0
620.98	Total Fumonisin, Other Rapid Test Kit (ppb)	2269	2,022	64.00	1,331	403.2	116.0	4			0
620.99	Total Fumonisin, Miscellaneous (ppb)	2302	852.0	0.0000				1			0
621.20	Fumonisin B1, LC (ppb)	2196	467.8	0.0000				1			0
621.20	Fumonisin B1, LC (ppb)	0563	< 100					1			5
621.21	Fumonisin B1, LC-FI OPA der. (ppb)	0098	< 200					0			5
621.23	Fumonisin B1, LC-MS/MS (ppb)	0009	184.4	0.3900	444.8	138.8	51.01	15	-1.88	29%	0
621.23	Fumonisin B1, LC-MS/MS (ppb)	0043	279.2	43.00	444.8	138.8	51.01	15	-1.19	19%	0
621.23	Fumonisin B1, LC-MS/MS (ppb)	0913	300.0	0.0000	444.8	138.8	51.01	15	-1.04	16%	0
621.23	Fumonisin B1, LC-MS/MS (ppb)	2053	307.6	8.000	444.8	138.8	51.01	15	-0.99	15%	0
621.23	Fumonisin B1, LC-MS/MS (ppb)	2207	350.0	100.0	444.8	138.8	51.01	15	-0.68	11%	0
621.23	Fumonisin B1, LC-MS/MS (ppb)	0964	408.5	6.800	444.8	138.8	51.01	15	-0.26	4%	0
621.23	Fumonisin B1, LC-MS/MS (ppb)	2136	425.0	54.00	444.8	138.8	51.01	15	-0.14	2%	0
621.23	Fumonisin B1, LC-MS/MS (ppb)	0010	475.0	170.0	444.8	138.8	51.01	15	0.22	3%	0
621.23	Fumonisin B1, LC-MS/MS (ppb)	2327	492.5	1.000	444.8	138.8	51.01	15	0.34	5%	0
621.23	Fumonisin B1, LC-MS/MS (ppb)	2140	529.5	43.00	444.8	138.8	51.01	15	0.61	10%	0
621.23	Fumonisin B1, LC-MS/MS (ppb)	0353	540.0	66.00	444.8	138.8	51.01	15	0.69	11%	0
621.23	Fumonisin B1, LC-MS/MS (ppb)	0870	550.0	20.00	444.8	138.8	51.01	15	0.76	12%	0
621.23	Fumonisin B1, LC-MS/MS (ppb)	0032	555.5	53.00	444.8	138.8	51.01	15	0.80	12%	0
621.23	Fumonisin B1, LC-MS/MS (ppb)	0682	570.0	0.0000	444.8	138.8	51.01	15	0.90	14%	0
621.23	Fumonisin B1, LC-MS/MS (ppb)	0160	700.0	200.0	444.8	138.8	51.01	15	1.84	29%	0
622.20	Fumonisin B2, LC (ppb)	2196	155.5	0.0000				1			0
622.21	Fumonisin B2, LC-FI OPA der. (ppb)	0098	< 200					0			5
622.23	Fumonisin B2, LC-MS/MS (ppb)	0009	79.30	0.9300	167.4	53.65	14.36	14	-1.64	26%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0913	100.0	0.0000	167.4	53.65	14.36	14	-1.26	20%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	2053	115.5	1.900	167.4	53.65	14.36	14	-0.97	16%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	2136	121.0	34.00	167.4	53.65	14.36	14	-0.86	14%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	2140	142.5	3.000	167.4	53.65	14.36	14	-0.46	7%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0043	147.5	6.600	167.4	53.65	14.36	14	-0.37	6%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0353	174.5	57.00	167.4	53.65	14.36	14	0.13	2%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0010	175.0	50.00	167.4	53.65	14.36	14	0.14	2%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0032	179.0	34.00	167.4	53.65	14.36	14	0.22	3%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	2327	186.0	10.00	167.4	53.65	14.36	14	0.35	6%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0870	200.0	0.0000	167.4	53.65	14.36	14	0.61	10%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0682	210.0	0.0000	167.4	53.65	14.36	14	0.79	13%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0964	254.4	3.587	167.4	53.65	14.36	14	1.62	26%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	0160	300.0	0.0000	167.4	53.65	14.36	14	2.47	40%	0
622.23	Fumonisin B2, LC-MS/MS (ppb)	2207	< 100		167.4	53.65	14.36	14			5
623.23	Fumonisin B3, LC-MS/MS (ppb)	0009	22.16	1.930	42.77	14.23	1.560	4			0
623.23	Fumonisin B3, LC-MS/MS (ppb)	2327	34.45	2.100	42.77	14.23	1.560	4			0
623.23	Fumonisin B3, LC-MS/MS (ppb)	2053	40.99	1.210	42.77	14.23	1.560	4			0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
623.23	Fumonisin B3, LC-MS/MS (ppb)	0353	73.50	1.000	42.77	14.23	1.560	4			0
623.23	Fumonisin B3, LC-MS/MS (ppb)	0032	49.70	10.00	42.77	14.23	1.560	4			1
623.23	Fumonisin B3, LC-MS/MS (ppb)	0010	< 56		42.77	14.23	1.560	4			5
623.23	Fumonisin B3, LC-MS/MS (ppb)	0160	< 100		42.77	14.23	1.560	4			5
623.23	Fumonisin B3, LC-MS/MS (ppb)	0870	< 100		42.77	14.23	1.560	4			5
623.23	Fumonisin B3, LC-MS/MS (ppb)	0913	< 100		42.77	14.23	1.560	4			5
630.01	Ochratoxin A, Neogen Veratox Ochratoxin (ppb)	0425	5.500	1.000				2			0
630.01	Ochratoxin A, Neogen Veratox Ochratoxin (ppb)	2098	5.830	1.400				2			0
630.03	Ochratoxin A, Romer AgraQuant Ochratoxin (ppb)	0723	17.56	0.1000				2			0
630.03	Ochratoxin A, Romer AgraQuant Ochratoxin (ppb)	2196	18.44	0.0000				2			0
630.04	Ochratoxin A, r-Biopharm Ridascreen Ochratoxin A (ppb)	2066	6.020	0.9600				1			0
630.05	Ochratoxin A, r-Biopharm Ridascreen FAST Och A (ppb)	2226	7.498	0.1630				3			0
630.05	Ochratoxin A, r-Biopharm Ridascreen FAST Och A (ppb)	0407	8.820	1.500				3			0
630.05	Ochratoxin A, r-Biopharm Ridascreen FAST Och A (ppb)	0066	14.63	2.070				3			0
630.06	Ochratoxin A, Vicam OchraTest (ppb)	2369	9.940	0.0400				2			0
630.06	Ochratoxin A, Vicam OchraTest (ppb)	0520	12.50	1.000				2			0
630.20	Ochratoxin A, LC (ppb)	2066	4.870	0.8200	12.18	4.190	1.258	10	-1.74	30%	0
630.20	Ochratoxin A, LC (ppb)	0723	6.125	0.0500	12.18	4.190	1.258	10	-1.44	25%	0
630.20	Ochratoxin A, LC (ppb)	0629	8.430	0.0200	12.18	4.190	1.258	10	-0.89	15%	0
630.20	Ochratoxin A, LC (ppb)	2369	9.940	0.0400	12.18	4.190	1.258	10	-0.53	9%	0
630.20	Ochratoxin A, LC (ppb)	2196	11.56	0.0000	12.18	4.190	1.258	10	-0.15	3%	0
630.20	Ochratoxin A, LC (ppb)	0563	12.87	0.2500	12.18	4.190	1.258	10	0.16	3%	0
630.20	Ochratoxin A, LC (ppb)	0674	15.37	0.0000	12.18	4.190	1.258	10	0.76	13%	0
630.20	Ochratoxin A, LC (ppb)	0051	15.95	8.900	12.18	4.190	1.258	10	0.90	15%	0
630.20	Ochratoxin A, LC (ppb)	0098	16.60	0.2000	12.18	4.190	1.258	10	1.06	18%	0
630.20	Ochratoxin A, LC (ppb)	2192	20.05	2.300	12.18	4.190	1.258	10	1.88	32%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	2136	7.385	1.170	12.36	4.253	1.194	14	-1.17	20%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0353	8.581	0.1920	12.36	4.253	1.194	14	-0.89	15%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	2207	10.50	1.000	12.36	4.253	1.194	14	-0.44	8%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	2129	10.65	1.100	12.36	4.253	1.194	14	-0.40	7%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0032	11.23	2.550	12.36	4.253	1.194	14	-0.27	5%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0964	11.48	3.680	12.36	4.253	1.194	14	-0.21	4%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	2140	11.50	1.000	12.36	4.253	1.194	14	-0.20	3%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0682	12.10	0.0000	12.36	4.253	1.194	14	-0.06	1%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0870	12.67	0.1000	12.36	4.253	1.194	14	0.07	1%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	2327	13.60	1.400	12.36	4.253	1.194	14	0.29	5%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0913	14.15	0.7000	12.36	4.253	1.194	14	0.42	7%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	2053	15.09	1.090	12.36	4.253	1.194	14	0.64	11%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0043	16.84	1.735	12.36	4.253	1.194	14	1.05	18%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0160	42.50	1.000	12.36	4.253	1.194	14	7.09	122%	0
630.23	Ochratoxin A, LC-MS/MS (ppb)	0010	< 1.5		12.36	4.253	1.194	14			5
630.98	Ochratoxin A, Other Rapid Test Kit (ppb)	2269	13.50	0.6000				1			0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
630.99	Ochratoxin A, Miscellaneous (ppb)	2302	181.5	7.000				1			0
640.03	T-2 Toxin, Romer AgraQuant T-2 (ppb)	0674	369.5	0.0000				3			0
640.03	T-2 Toxin, Romer AgraQuant T-2 (ppb)	0066	375.3	36.48				3			0
640.03	T-2 Toxin, Romer AgraQuant T-2 (ppb)	2196	423.8	0.0000				3			0
640.05	T-2 Toxin, r-Biopharm Ridascreen FAST T-2 (ppb)	0407	295.9	13.05				1			0
640.23	T-2 Toxin, LC-MS/MS (ppb)	0682	240.0	0.0000	313.6	98.83	28.99	12	-0.75	12%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	2140	264.5	29.00	313.6	98.83	28.99	12	-0.50	8%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	2207	265.0	30.00	313.6	98.83	28.99	12	-0.49	8%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	2136	269.5	9.000	313.6	98.83	28.99	12	-0.45	7%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	0913	271.5	21.50	313.6	98.83	28.99	12	-0.43	7%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	2327	289.5	9.000	313.6	98.83	28.99	12	-0.24	4%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	0870	304.2	7.140	313.6	98.83	28.99	12	-0.10	2%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	0353	322.5	17.00	313.6	98.83	28.99	12	0.09	1%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	2129	352.5	41.00	313.6	98.83	28.99	12	0.39	6%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	0964	357.0	18.70	313.6	98.83	28.99	12	0.44	7%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	2053	415.0	35.50	313.6	98.83	28.99	12	1.03	16%	0
640.23	T-2 Toxin, LC-MS/MS (ppb)	0160	445.0	130.0	313.6	98.83	28.99	12	1.33	21%	0
640.25	T-2 Toxin, GC-MS (ppb)	0563	283.5	19.00				1			0
640.98	T-2 Toxin, Other Rapid Test Kit (ppb)	0297	276.0	10.00				2			0
640.98	T-2 Toxin, Other Rapid Test Kit (ppb)	2269	2,353	117.4				2			0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	0964	70.62	12.51	89.25	29.10	9.277	9	-0.64	10%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	2327	75.15	10.90	89.25	29.10	9.277	9	-0.48	8%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	0870	78.91	0.7800	89.25	29.10	9.277	9	-0.36	6%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	2207	85.00	10.00	89.25	29.10	9.277	9	-0.15	2%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	0353	90.50	33.00	89.25	29.10	9.277	9	0.04	1%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	2140	92.00	4.000	89.25	29.10	9.277	9	0.09	2%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	2136	99.15	3.700	89.25	29.10	9.277	9	0.34	6%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	0913	100.2	3.500	89.25	29.10	9.277	9	0.37	6%	0
641.23	HT-2 Toxin, LC-MS/MS (ppb)	2053	123.7	5.100	89.25	29.10	9.277	9	1.18	19%	0
642.04	T-2 Toxin + HT-2 Toxin, LC-MS (ppb)	0353	479.5	81.00				1			0
642.21	T-2 Toxin + HT-2 Toxin, Neogen Veratox T-2 + HT-2 (ppb)	0425	195.0	2.000	588.3	182.2	37.50	4			0
642.21	T-2 Toxin + HT-2 Toxin, Neogen Veratox T-2 + HT-2 (ppb)	0227	463.5	15.00	588.3	182.2	37.50	4			0
642.21	T-2 Toxin + HT-2 Toxin, Neogen Veratox T-2 + HT-2 (ppb)	2192	653.0	54.00	588.3	182.2	37.50	4			0
642.21	T-2 Toxin + HT-2 Toxin, Neogen Veratox T-2 + HT-2 (ppb)	2098	1,042	79.00	588.3	182.2	37.50	4			0
642.98	T-2 Toxin + HT-2 Toxin, Other Rapid Test Kit (ppb)	0202	238.0	8.000				3			0
642.98	T-2 Toxin + HT-2 Toxin, Other Rapid Test Kit (ppb)	0142	268.4	37.10				3			0
642.98	T-2 Toxin + HT-2 Toxin, Other Rapid Test Kit (ppb)	0590	327.0	38.00				3			0
642.98	T-2 Toxin + HT-2 Toxin, Other Rapid Test Kit (ppb)	2269	3,246	162.1				3			2
642.99	T-2 Toxin + HT-2 Toxin, Miscellaneous Technique (ppb)	2302	244.5	3.000				2			0
642.99	T-2 Toxin + HT-2 Toxin, Miscellaneous Technique (ppb)	0407	268.8	29.86				2			0
650.01	Zearalenone, Neogen Veratox Zearalenone (ppb)	2113	110.0	0.0000	155.5	49.95	5.670	6	-0.91	15%	0
650.01	Zearalenone, Neogen Veratox Zearalenone (ppb)	0027	122.5	19.00	155.5	49.95	5.670	6	-0.66	11%	0



Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			
650.01	Zearalenone, Neogen Veratox Zearalenone (ppb)	0425	148.0	4.000	155.5	49.95	5.670	6	-0.15	2%	0
650.01	Zearalenone, Neogen Veratox Zearalenone (ppb)	0870	153.1	7.018	155.5	49.95	5.670	6	-0.05	1%	0
650.01	Zearalenone, Neogen Veratox Zearalenone (ppb)	0297	174.0	2.000	155.5	49.95	5.670	6	0.37	6%	0
650.01	Zearalenone, Neogen Veratox Zearalenone (ppb)	2098	236.7	2.000	155.5	49.95	5.670	6	1.63	26%	0
650.04	Zearalenone, Romer AgraQuant ZON (ppb)	0674	143.0	0.0000				3			0
650.04	Zearalenone, Romer AgraQuant ZON (ppb)	2196	168.6	0.0000				3			0
650.04	Zearalenone, Romer AgraQuant ZON (ppb)	0723	221.9	0.0900				3			0
650.06	Zearalenone, r-Biopharm Ridascreen FAST ZON (ppb)	2226	114.3	17.43	123.6	39.94	24.36	4			0
650.06	Zearalenone, r-Biopharm Ridascreen FAST ZON (ppb)	0202	119.5	17.00	123.6	39.94	24.36	4			0
650.06	Zearalenone, r-Biopharm Ridascreen FAST ZON (ppb)	2066	124.5	33.00	123.6	39.94	24.36	4			0
650.06	Zearalenone, r-Biopharm Ridascreen FAST ZON (ppb)	0590	136.0	30.00	123.6	39.94	24.36	4			0
650.07	Zearalenone, r-Biopharm Ridascreen FAST ZON SC (ppb)	0407	114.5	1.200				3			0
650.07	Zearalenone, r-Biopharm Ridascreen FAST ZON SC (ppb)	0066	134.0	15.50				3			0
650.07	Zearalenone, r-Biopharm Ridascreen FAST ZON SC (ppb)	0035	159.1	5.000				3			0
650.20	Zearalenone, LC (ppb)	0723	77.56	17.46	177.9	56.94	27.24	7	-1.76	28%	0
650.20	Zearalenone, LC (ppb)	2369	134.7	12.51	177.9	56.94	27.24	7	-0.76	12%	0
650.20	Zearalenone, LC (ppb)	2196	161.4	0.0000	177.9	56.94	27.24	7	-0.29	5%	0
650.20	Zearalenone, LC (ppb)	2192	168.5	3.000	177.9	56.94	27.24	7	-0.17	3%	0
650.20	Zearalenone, LC (ppb)	0098	195.0	50.00	177.9	56.94	27.24	7	0.30	5%	0
650.20	Zearalenone, LC (ppb)	0563	216.0	4.000	177.9	56.94	27.24	7	0.67	11%	0
650.20	Zearalenone, LC (ppb)	0646	557.8	103.7	177.9	56.94	27.24	7	6.67	107%	0
650.21	Zearalenone, LC-FI, ISO (ppb)	2369	134.7	12.51				2			0
650.21	Zearalenone, LC-FI, ISO (ppb)	0051	200.0	2.000				2			0
650.23	Zearalenone, LC-MS (ppb)	0353	129.5	1.000				1			0
650.24	Zearalenone, LC-MS/MS (ppb)	2207	115.0	30.00	183.6	58.69	33.44	13	-1.17	19%	0
650.24	Zearalenone, LC-MS/MS (ppb)	2136	133.5	45.00	183.6	58.69	33.44	13	-0.85	14%	0
650.24	Zearalenone, LC-MS/MS (ppb)	2129	138.0	32.00	183.6	58.69	33.44	13	-0.78	12%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0913	138.1	10.90	183.6	58.69	33.44	13	-0.78	12%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0682	151.8	0.0000	183.6	58.69	33.44	13	-0.54	9%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0032	153.0	8.000	183.6	58.69	33.44	13	-0.52	8%	0
650.24	Zearalenone, LC-MS/MS (ppb)	2140	178.0	20.00	183.6	58.69	33.44	13	-0.09	2%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0010	200.0	0.0000	183.6	58.69	33.44	13	0.28	4%	0
650.24	Zearalenone, LC-MS/MS (ppb)	2327	202.5	17.00	183.6	58.69	33.44	13	0.32	5%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0160	231.0	6.000	183.6	58.69	33.44	13	0.81	13%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0964	233.6	17.60	183.6	58.69	33.44	13	0.85	14%	0
650.24	Zearalenone, LC-MS/MS (ppb)	2053	244.0	21.80	183.6	58.69	33.44	13	1.03	16%	0
650.24	Zearalenone, LC-MS/MS (ppb)	0043	422.4	226.4	183.6	58.69	33.44	13	4.07	65%	0
650.98	Zearalenone, Other Rapid Test Kit (ppb)	0142	104.9	20.90				3			0
650.98	Zearalenone, Other Rapid Test Kit (ppb)	2269	222.8	28.20				3			0
650.98	Zearalenone, Other Rapid Test Kit (ppb)	2050	352.1	40.20				3			0
650.99	Zearalenone, Miscellaneous (ppb)	0520	0.0820	0.0160				2			0
650.99	Zearalenone, Miscellaneous (ppb)	2302	320.0	20.00				2			0

Method Code	Analyte Name and Method (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Threshold %RSD	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Labs			

Notes: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2 and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = Rejected for duplicates too far apart, 2 = Rejected as outlier, 5 = A reporting limit and 4 = zeros submitted as values. Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). Flag 3 indicates not used in statistics.