

**AAFCO Check Sample Program**  
**All Labs and All Methods Report**  
**Sort by Method**  
**Proficiency Testing**  
**Sample # 201424**  
**Pelleted Beef Feed, Medicated**



Robust statistics not used if < 6 labs reporting, in this case the Z Scores are included for information only (Grey).

Issue Date : 05/31/2014

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
000	Urea (%)	0920	0.00500	0.01000	0.20250	0.27931	0.10500	2	-0.71	000.99	0
000	Urea (%)	0278	0.40000	0.20000	0.20250	0.27931	0.10500	2	0.71	000.02	0
001	Loss on Drying (%)	0618	2.8850	0.21000	7.0636	0.26289	0.09628	75	-15.89	001.07	0
001	Loss on Drying (%)	0618	5.4400	0.02000	7.0636	0.26289	0.09628	75	-6.18	001.99	0
001	Loss on Drying (%)	0675	5.5300	0.12000	7.0636	0.26289	0.09628	75	-5.83	001.07	0
001	Loss on Drying (%)	2048	6.1400	0.12000	7.0636	0.26289	0.09628	75	-3.51	001.00	0
001	Loss on Drying (%)	0536	6.1525	0.25500	7.0636	0.26289	0.09628	75	-3.47	001.99	0
001	Loss on Drying (%)	0015	6.3100	0.26000	7.0636	0.26289	0.09628	75	-2.87	001.07	0
001	Loss on Drying (%)	0345	6.3500	0.06000	7.0636	0.26289	0.09628	75	-2.71	001.00	0
001	Loss on Drying (%)	0038	6.4600	0.00000	7.0636	0.26289	0.09628	75	-2.30	001.07	0
001	Loss on Drying (%)	0720	6.6600	0.22000	7.0636	0.26289	0.09628	75	-1.54	001.99	0
001	Loss on Drying (%)	0045	6.6850	0.49000	7.0636	0.26289	0.09628	75	-1.44	001.07	0
001	Loss on Drying (%)	0297	6.7400	0.02000	7.0636	0.26289	0.09628	75	-1.23	001.07	0
001	Loss on Drying (%)	0309	6.7495	0.11100	7.0636	0.26289	0.09628	75	-1.19	001.00	0
001	Loss on Drying (%)	0171	6.8100	0.02000	7.0636	0.26289	0.09628	75	-0.96	001.07	0
001	Loss on Drying (%)	0353	6.8250	0.07000	7.0636	0.26289	0.09628	75	-0.91	001.07	0
001	Loss on Drying (%)	0083	6.8400	0.06000	7.0636	0.26289	0.09628	75	-0.85	001.07	0
001	Loss on Drying (%)	0178	6.8450	0.11000	7.0636	0.26289	0.09628	75	-0.83	001.07	0
001	Loss on Drying (%)	0891	6.8528	0.00170	7.0636	0.26289	0.09628	75	-0.80	001.03	0
001	Loss on Drying (%)	0918	6.8550	0.39000	7.0636	0.26289	0.09628	75	-0.79	001.99	0
001	Loss on Drying (%)	2018	6.8750	0.03000	7.0636	0.26289	0.09628	75	-0.72	001.99	0
001	Loss on Drying (%)	0845	6.8900	0.14000	7.0636	0.26289	0.09628	75	-0.66	001.07	0
001	Loss on Drying (%)	2058	6.9000	0.14000	7.0636	0.26289	0.09628	75	-0.62	001.99	0
001	Loss on Drying (%)	0089	6.9600	0.00000	7.0636	0.26289	0.09628	75	-0.39	001.07	0
001	Loss on Drying (%)	0098	6.9650	0.01000	7.0636	0.26289	0.09628	75	-0.37	001.07	0
001	Loss on Drying (%)	0278	6.9700	0.46000	7.0636	0.26289	0.09628	75	-0.36	001.07	0
001	Loss on Drying (%)	0035	6.9900	0.12000	7.0636	0.26289	0.09628	75	-0.28	001.07	0
001	Loss on Drying (%)	0049	6.9950	0.47000	7.0636	0.26289	0.09628	75	-0.26	001.07	0
001	Loss on Drying (%)	0948	6.9950	0.01000	7.0636	0.26289	0.09628	75	-0.26	001.99	0
001	Loss on Drying (%)	0610	7.0000	0.20000	7.0636	0.26289	0.09628	75	-0.24	001.05	0
001	Loss on Drying (%)	0510	7.0000	0.00000	7.0636	0.26289	0.09628	75	-0.24	001.99	0
001	Loss on Drying (%)	0638	7.0300	0.02000	7.0636	0.26289	0.09628	75	-0.13	001.07	0
001	Loss on Drying (%)	0357	7.0450	0.05000	7.0636	0.26289	0.09628	75	-0.07	001.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
001	Loss on Drying (%)	0693	7.0500	0.34000	7.0636	0.26289	0.09628	75	-0.05	001.07	0
001	Loss on Drying (%)	0676	7.0500	0.10000	7.0636	0.26289	0.09628	75	-0.05	001.99	0
001	Loss on Drying (%)	2059	7.0550	0.05000	7.0636	0.26289	0.09628	75	-0.03	001.07	0
001	Loss on Drying (%)	0065	7.0600	0.00000	7.0636	0.26289	0.09628	75	-0.01	001.07	0
001	Loss on Drying (%)	2042	7.0750	0.01000	7.0636	0.26289	0.09628	75	0.04	001.03	0
001	Loss on Drying (%)	0843	7.0850	0.03000	7.0636	0.26289	0.09628	75	0.08	001.07	0
001	Loss on Drying (%)	0951	7.0850	0.01000	7.0636	0.26289	0.09628	75	0.08	001.99	0
001	Loss on Drying (%)	0337	7.0900	0.14000	7.0636	0.26289	0.09628	75	0.10	001.08	0
001	Loss on Drying (%)	0868	7.0950	0.03000	7.0636	0.26289	0.09628	75	0.12	001.03	0
001	Loss on Drying (%)	0226	7.1000	0.00000	7.0636	0.26289	0.09628	75	0.14	001.07	0
001	Loss on Drying (%)	0689	7.1000	0.00000	7.0636	0.26289	0.09628	75	0.14	001.07	0
001	Loss on Drying (%)	0187	7.1200	0.04000	7.0636	0.26289	0.09628	75	0.21	001.07	0
001	Loss on Drying (%)	2063	7.1250	0.01000	7.0636	0.26289	0.09628	75	0.23	001.03	0
001	Loss on Drying (%)	0849	7.1275	0.09500	7.0636	0.26289	0.09628	75	0.24	001.07	0
001	Loss on Drying (%)	0571	7.1300	0.00000	7.0636	0.26289	0.09628	75	0.25	001.07	0
001	Loss on Drying (%)	0619	7.1400	0.00000	7.0636	0.26289	0.09628	75	0.29	001.03	0
001	Loss on Drying (%)	0895	7.1450	0.01000	7.0636	0.26289	0.09628	75	0.31	001.03	0
001	Loss on Drying (%)	0889	7.1500	0.04000	7.0636	0.26289	0.09628	75	0.33	001.99	0
001	Loss on Drying (%)	0878	7.1550	0.03000	7.0636	0.26289	0.09628	75	0.35	001.03	0
001	Loss on Drying (%)	0893	7.1550	0.01000	7.0636	0.26289	0.09628	75	0.35	001.03	0
001	Loss on Drying (%)	0139	7.1700	0.00000	7.0636	0.26289	0.09628	75	0.40	001.07	0
001	Loss on Drying (%)	0004	7.1700	0.04000	7.0636	0.26289	0.09628	75	0.40	001.99	0
001	Loss on Drying (%)	0590	7.1750	0.29000	7.0636	0.26289	0.09628	75	0.42	001.08	0
001	Loss on Drying (%)	0903	7.1950	0.01000	7.0636	0.26289	0.09628	75	0.50	001.03	0
001	Loss on Drying (%)	0937	7.2000	0.00000	7.0636	0.26289	0.09628	75	0.52	001.03	0
001	Loss on Drying (%)	2065	7.2317	0.01000	7.0636	0.26289	0.09628	75	0.64	001.03	0
001	Loss on Drying (%)	0504	7.2350	0.45000	7.0636	0.26289	0.09628	75	0.65	001.00	0
001	Loss on Drying (%)	2062	7.2643	0.03860	7.0636	0.26289	0.09628	75	0.76	001.03	0
001	Loss on Drying (%)	0683	7.2750	0.11000	7.0636	0.26289	0.09628	75	0.80	001.07	0
001	Loss on Drying (%)	0629	7.2800	0.02000	7.0636	0.26289	0.09628	75	0.82	001.99	0
001	Loss on Drying (%)	2012	7.2950	0.01000	7.0636	0.26289	0.09628	75	0.88	001.07	0
001	Loss on Drying (%)	0505	7.2950	0.03000	7.0636	0.26289	0.09628	75	0.88	001.99	0
001	Loss on Drying (%)	0413	7.3000	0.20000	7.0636	0.26289	0.09628	75	0.90	001.07	0
001	Loss on Drying (%)	0652	7.3000	0.20000	7.0636	0.26289	0.09628	75	0.90	001.08	0
001	Loss on Drying (%)	0897	7.3400	0.02000	7.0636	0.26289	0.09628	75	1.05	001.03	0
001	Loss on Drying (%)	0872	7.3700	0.06000	7.0636	0.26289	0.09628	75	1.17	001.07	0
001	Loss on Drying (%)	0581	7.3800	0.14000	7.0636	0.26289	0.09628	75	1.20	001.07	0
001	Loss on Drying (%)	0142	7.4650	0.01000	7.0636	0.26289	0.09628	75	1.53	001.07	0
001	Loss on Drying (%)	0609	7.5000	0.20000	7.0636	0.26289	0.09628	75	1.66	001.07	0
001	Loss on Drying (%)	0037	7.6850	0.05000	7.0636	0.26289	0.09628	75	2.36	001.99	0
001	Loss on Drying (%)	0695	7.6950	0.17000	7.0636	0.26289	0.09628	75	2.40	001.07	0
001	Loss on Drying (%)	0852	7.8000	0.00000	7.0636	0.26289	0.09628	75	2.80	001.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
001	Loss on Drying (%)	2068	8.9400	0.00000	7.0636	0.26289	0.09628	75	7.14	001.00	0
001	Loss on Drying (%)	2011	10.115	0.03000	7.0636	0.26289	0.09628	75	11.61	001.07	0
001	Loss on Drying (%)	0074	6.9250	0.65000	7.0636	0.26289	0.09628	75	-0.53	001.07	1
002	Protein (%)	0951	13.815	0.11000	15.458	0.25981	0.09704	210	-6.32	002.11	0
002	Protein (%)	0720	14.595	0.03000	15.458	0.25981	0.09704	210	-3.32	002.11	0
002	Protein (%)	0042	14.710	0.04000	15.458	0.25981	0.09704	210	-2.88	002.02	0
002	Protein (%)	0889	14.785	0.35000	15.458	0.25981	0.09704	210	-2.59	002.11	0
002	Protein (%)	0885	14.860	0.02000	15.458	0.25981	0.09704	210	-2.30	002.05	0
002	Protein (%)	0728	14.880	0.14000	15.458	0.25981	0.09704	210	-2.22	002.04	0
002	Protein (%)	0596	14.940	0.40000	15.458	0.25981	0.09704	210	-1.99	002.05	0
002	Protein (%)	0676	14.950	0.30000	15.458	0.25981	0.09704	210	-1.95	002.06	0
002	Protein (%)	0921	14.970	0.08000	15.458	0.25981	0.09704	210	-1.88	002.05	0
002	Protein (%)	0187	14.980	0.02000	15.458	0.25981	0.09704	210	-1.84	002.04	0
002	Protein (%)	0510	15.000	0.00000	15.458	0.25981	0.09704	210	-1.76	002.06	0
002	Protein (%)	0042	15.040	0.12000	15.458	0.25981	0.09704	210	-1.61	002.06	0
002	Protein (%)	0199	15.050	0.08000	15.458	0.25981	0.09704	210	-1.57	002.01	0
002	Protein (%)	0536	15.070	0.26000	15.458	0.25981	0.09704	210	-1.49	002.05	0
002	Protein (%)	0144	15.075	0.09000	15.458	0.25981	0.09704	210	-1.47	002.06	0
002	Protein (%)	0921	15.075	0.41000	15.458	0.25981	0.09704	210	-1.47	002.06	0
002	Protein (%)	0083	15.100	0.20000	15.458	0.25981	0.09704	210	-1.38	002.06	0
002	Protein (%)	2048	15.110	0.12000	15.458	0.25981	0.09704	210	-1.34	002.05	0
002	Protein (%)	0848	15.115	0.11000	15.458	0.25981	0.09704	210	-1.32	002.01	0
002	Protein (%)	0043	15.120	0.02000	15.458	0.25981	0.09704	210	-1.30	002.01	0
002	Protein (%)	2022	15.120	0.24000	15.458	0.25981	0.09704	210	-1.30	002.01	0
002	Protein (%)	0874	15.140	0.22000	15.458	0.25981	0.09704	210	-1.22	002.04	0
002	Protein (%)	0955	15.150	0.10000	15.458	0.25981	0.09704	210	-1.18	002.06	0
002	Protein (%)	0242	15.150	0.20000	15.458	0.25981	0.09704	210	-1.18	002.06	0
002	Protein (%)	0941	15.155	0.17000	15.458	0.25981	0.09704	210	-1.17	002.05	0
002	Protein (%)	0618	15.155	0.03000	15.458	0.25981	0.09704	210	-1.17	002.06	0
002	Protein (%)	0937	15.165	0.05000	15.458	0.25981	0.09704	210	-1.13	002.05	0
002	Protein (%)	0505	15.165	0.03000	15.458	0.25981	0.09704	210	-1.13	002.06	0
002	Protein (%)	0620	15.169	0.11340	15.458	0.25981	0.09704	210	-1.11	002.05	0
002	Protein (%)	0939	15.180	0.44000	15.458	0.25981	0.09704	210	-1.07	002.00	0
002	Protein (%)	0685	15.185	0.15000	15.458	0.25981	0.09704	210	-1.05	002.01	0
002	Protein (%)	0405	15.190	0.16000	15.458	0.25981	0.09704	210	-1.03	002.04	0
002	Protein (%)	0138	15.199	0.27000	15.458	0.25981	0.09704	210	-1.00	002.06	0
002	Protein (%)	0968	15.200	0.01100	15.458	0.25981	0.09704	210	-0.99	002.01	0
002	Protein (%)	0226	15.200	0.20000	15.458	0.25981	0.09704	210	-0.99	002.06	0
002	Protein (%)	0208	15.200	0.00000	15.458	0.25981	0.09704	210	-0.99	002.08	0
002	Protein (%)	0651	15.216	0.21200	15.458	0.25981	0.09704	210	-0.93	002.05	0
002	Protein (%)	0723	15.219	0.00000	15.458	0.25981	0.09704	210	-0.92	002.01	0
002	Protein (%)	0337	15.235	0.33000	15.458	0.25981	0.09704	210	-0.86	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
002	Protein (%)	2012	15.235	0.01000	15.458	0.25981	0.09704	210	-0.86	002.06	0
002	Protein (%)	0028	15.240	0.08000	15.458	0.25981	0.09704	210	-0.84	002.00	0
002	Protein (%)	0952	15.245	0.05000	15.458	0.25981	0.09704	210	-0.82	002.05	0
002	Protein (%)	0675	15.250	0.06000	15.458	0.25981	0.09704	210	-0.80	002.05	0
002	Protein (%)	0897	15.250	0.10000	15.458	0.25981	0.09704	210	-0.80	002.05	0
002	Protein (%)	0357	15.250	0.00000	15.458	0.25981	0.09704	210	-0.80	002.06	0
002	Protein (%)	0726	15.251	0.00900	15.458	0.25981	0.09704	210	-0.80	002.06	0
002	Protein (%)	0354	15.255	0.03000	15.458	0.25981	0.09704	210	-0.78	002.05	0
002	Protein (%)	0951	15.255	0.07000	15.458	0.25981	0.09704	210	-0.78	002.05	0
002	Protein (%)	2062	15.255	0.01500	15.458	0.25981	0.09704	210	-0.78	002.05	0
002	Protein (%)	0036	15.258	0.02390	15.458	0.25981	0.09704	210	-0.77	002.02	0
002	Protein (%)	2058	15.265	0.03000	15.458	0.25981	0.09704	210	-0.74	002.05	0
002	Protein (%)	0263	15.268	0.02460	15.458	0.25981	0.09704	210	-0.73	002.06	0
002	Protein (%)	0643	15.280	0.04000	15.458	0.25981	0.09704	210	-0.68	002.99	0
002	Protein (%)	2068	15.285	0.13000	15.458	0.25981	0.09704	210	-0.66	002.01	0
002	Protein (%)	0039	15.289	0.16760	15.458	0.25981	0.09704	210	-0.65	002.05	0
002	Protein (%)	0720	15.290	0.04000	15.458	0.25981	0.09704	210	-0.65	002.06	0
002	Protein (%)	0202	15.295	0.17000	15.458	0.25981	0.09704	210	-0.63	002.06	0
002	Protein (%)	0619	15.300	0.00000	15.458	0.25981	0.09704	210	-0.61	002.05	0
002	Protein (%)	0278	15.300	0.00000	15.458	0.25981	0.09704	210	-0.61	002.06	0
002	Protein (%)	0062	15.305	0.21600	15.458	0.25981	0.09704	210	-0.59	002.01	0
002	Protein (%)	0353	15.305	0.07000	15.458	0.25981	0.09704	210	-0.59	002.06	0
002	Protein (%)	0960	15.305	0.29000	15.458	0.25981	0.09704	210	-0.59	002.06	0
002	Protein (%)	2065	15.312	0.01000	15.458	0.25981	0.09704	210	-0.56	002.05	0
002	Protein (%)	0948	15.315	0.05000	15.458	0.25981	0.09704	210	-0.55	002.05	0
002	Protein (%)	2059	15.316	0.28130	15.458	0.25981	0.09704	210	-0.55	002.06	0
002	Protein (%)	2011	15.320	0.04000	15.458	0.25981	0.09704	210	-0.53	002.05	0
002	Protein (%)	0626	15.327	0.05600	15.458	0.25981	0.09704	210	-0.50	002.06	0
002	Protein (%)	0008	15.330	0.10000	15.458	0.25981	0.09704	210	-0.49	002.06	0
002	Protein (%)	0878	15.335	0.03000	15.458	0.25981	0.09704	210	-0.47	002.05	0
002	Protein (%)	2006	15.335	0.07000	15.458	0.25981	0.09704	210	-0.47	002.05	0
002	Protein (%)	0953	15.340	0.08000	15.458	0.25981	0.09704	210	-0.45	002.06	0
002	Protein (%)	2006	15.340	0.04000	15.458	0.25981	0.09704	210	-0.45	002.06	0
002	Protein (%)	0683	15.345	0.09000	15.458	0.25981	0.09704	210	-0.43	002.05	0
002	Protein (%)	2042	15.350	0.10000	15.458	0.25981	0.09704	210	-0.41	002.05	0
002	Protein (%)	0017	15.350	0.10000	15.458	0.25981	0.09704	210	-0.41	002.06	0
002	Protein (%)	0194	15.355	0.03000	15.458	0.25981	0.09704	210	-0.40	002.05	0
002	Protein (%)	0881	15.356	0.06600	15.458	0.25981	0.09704	210	-0.39	002.01	0
002	Protein (%)	0849	15.359	0.03800	15.458	0.25981	0.09704	210	-0.38	002.05	0
002	Protein (%)	0868	15.360	0.02000	15.458	0.25981	0.09704	210	-0.38	002.05	0
002	Protein (%)	0695	15.370	0.10000	15.458	0.25981	0.09704	210	-0.34	002.06	0
002	Protein (%)	0208	15.375	0.03000	15.458	0.25981	0.09704	210	-0.32	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
002	Protein (%)	0066	15.380	0.08000	15.458	0.25981	0.09704	210	-0.30	002.02	0
002	Protein (%)	0552	15.390	0.02000	15.458	0.25981	0.09704	210	-0.26	002.05	0
002	Protein (%)	0674	15.390	0.00000	15.458	0.25981	0.09704	210	-0.26	002.05	0
002	Protein (%)	0407	15.400	0.10000	15.458	0.25981	0.09704	210	-0.22	002.06	0
002	Protein (%)	0852	15.400	0.20000	15.458	0.25981	0.09704	210	-0.22	002.05	0
002	Protein (%)	0047	15.400	0.00000	15.458	0.25981	0.09704	210	-0.22	002.06	0
002	Protein (%)	0051	15.400	0.00000	15.458	0.25981	0.09704	210	-0.22	002.06	0
002	Protein (%)	0139	15.415	0.01000	15.458	0.25981	0.09704	210	-0.16	002.06	0
002	Protein (%)	0571	15.420	0.02200	15.458	0.25981	0.09704	210	-0.15	002.06	0
002	Protein (%)	0504	15.430	0.10000	15.458	0.25981	0.09704	210	-0.11	002.04	0
002	Protein (%)	0037	15.430	0.04000	15.458	0.25981	0.09704	210	-0.11	002.06	0
002	Protein (%)	0043	15.435	0.21000	15.458	0.25981	0.09704	210	-0.09	002.06	0
002	Protein (%)	0100	15.440	0.00000	15.458	0.25981	0.09704	210	-0.07	002.06	0
002	Protein (%)	0098	15.440	0.20000	15.458	0.25981	0.09704	210	-0.07	002.08	0
002	Protein (%)	2012	15.440	0.08000	15.458	0.25981	0.09704	210	-0.07	002.05	0
002	Protein (%)	0598	15.445	0.31000	15.458	0.25981	0.09704	210	-0.05	002.06	0
002	Protein (%)	0652	15.450	0.10000	15.458	0.25981	0.09704	210	-0.03	002.01	0
002	Protein (%)	0638	15.450	0.00000	15.458	0.25981	0.09704	210	-0.03	002.04	0
002	Protein (%)	0354	15.450	0.04000	15.458	0.25981	0.09704	210	-0.03	002.06	0
002	Protein (%)	0619	15.450	0.10000	15.458	0.25981	0.09704	210	-0.03	002.06	0
002	Protein (%)	0893	15.455	0.01000	15.458	0.25981	0.09704	210	-0.01	002.05	0
002	Protein (%)	0038	15.455	0.07000	15.458	0.25981	0.09704	210	-0.01	002.06	0
002	Protein (%)	0868	15.460	0.04000	15.458	0.25981	0.09704	210	0.01	002.06	0
002	Protein (%)	0297	15.470	0.12000	15.458	0.25981	0.09704	210	0.05	002.06	0
002	Protein (%)	0199	15.470	0.02000	15.458	0.25981	0.09704	210	0.05	002.06	0
002	Protein (%)	0074	15.475	0.07000	15.458	0.25981	0.09704	210	0.07	002.06	0
002	Protein (%)	0590	15.475	0.05000	15.458	0.25981	0.09704	210	0.07	002.06	0
002	Protein (%)	0891	15.481	0.00500	15.458	0.25981	0.09704	210	0.09	002.05	0
002	Protein (%)	0089	15.490	0.00000	15.458	0.25981	0.09704	210	0.12	002.06	0
002	Protein (%)	0861	15.490	0.04000	15.458	0.25981	0.09704	210	0.12	002.06	0
002	Protein (%)	0948	15.490	0.14000	15.458	0.25981	0.09704	210	0.12	002.06	0
002	Protein (%)	2023	15.490	0.02000	15.458	0.25981	0.09704	210	0.12	002.06	0
002	Protein (%)	0065	15.496	0.04630	15.458	0.25981	0.09704	210	0.15	002.06	0
002	Protein (%)	0689	15.500	0.20000	15.458	0.25981	0.09704	210	0.16	002.05	0
002	Protein (%)	0010	15.500	0.20000	15.458	0.25981	0.09704	210	0.16	002.06	0
002	Protein (%)	0171	15.500	0.20000	15.458	0.25981	0.09704	210	0.16	002.06	0
002	Protein (%)	0229	15.500	0.04000	15.458	0.25981	0.09704	210	0.16	002.06	0
002	Protein (%)	0910	15.500	0.20000	15.458	0.25981	0.09704	210	0.16	002.06	0
002	Protein (%)	0009	15.505	0.07000	15.458	0.25981	0.09704	210	0.18	002.06	0
002	Protein (%)	0205	15.505	0.01000	15.458	0.25981	0.09704	210	0.18	002.06	0
002	Protein (%)	0913	15.505	0.09000	15.458	0.25981	0.09704	210	0.18	002.06	0
002	Protein (%)	0148	15.515	0.05000	15.458	0.25981	0.09704	210	0.22	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
002	Protein (%)	0674	15.515	0.01000	15.458	0.25981	0.09704	210	0.22	002.06	0
002	Protein (%)	0857	15.515	0.01000	15.458	0.25981	0.09704	210	0.22	002.06	0
002	Protein (%)	0298	15.520	0.02000	15.458	0.25981	0.09704	210	0.24	002.06	0
002	Protein (%)	2003	15.520	0.00000	15.458	0.25981	0.09704	210	0.24	002.06	0
002	Protein (%)	0035	15.525	0.01000	15.458	0.25981	0.09704	210	0.26	002.06	0
002	Protein (%)	0309	15.525	0.05000	15.458	0.25981	0.09704	210	0.26	002.06	0
002	Protein (%)	0004	15.530	0.02000	15.458	0.25981	0.09704	210	0.28	002.06	0
002	Protein (%)	2023	15.530	0.10000	15.458	0.25981	0.09704	210	0.28	002.01	0
002	Protein (%)	0682	15.535	0.03000	15.458	0.25981	0.09704	210	0.30	002.06	0
002	Protein (%)	0300	15.540	0.38000	15.458	0.25981	0.09704	210	0.32	002.06	0
002	Protein (%)	2018	15.540	0.08000	15.458	0.25981	0.09704	210	0.32	002.06	0
002	Protein (%)	0955	15.550	0.10000	15.458	0.25981	0.09704	210	0.36	002.05	0
002	Protein (%)	0175	15.550	0.10000	15.458	0.25981	0.09704	210	0.36	002.06	0
002	Protein (%)	0413	15.550	0.10000	15.458	0.25981	0.09704	210	0.36	002.06	0
002	Protein (%)	2004	15.550	0.10000	15.458	0.25981	0.09704	210	0.36	002.99	0
002	Protein (%)	2058	15.555	0.13000	15.458	0.25981	0.09704	210	0.37	002.06	0
002	Protein (%)	0168	15.560	0.12000	15.458	0.25981	0.09704	210	0.39	002.06	0
002	Protein (%)	2063	15.560	0.02000	15.458	0.25981	0.09704	210	0.39	002.05	0
002	Protein (%)	0939	15.565	0.17000	15.458	0.25981	0.09704	210	0.41	002.02	0
002	Protein (%)	0903	15.565	0.01000	15.458	0.25981	0.09704	210	0.41	002.05	0
002	Protein (%)	0164	15.570	0.00000	15.458	0.25981	0.09704	210	0.43	002.06	0
002	Protein (%)	0693	15.570	0.26000	15.458	0.25981	0.09704	210	0.43	002.06	0
002	Protein (%)	0015	15.575	0.03000	15.458	0.25981	0.09704	210	0.45	002.05	0
002	Protein (%)	0190	15.580	0.02000	15.458	0.25981	0.09704	210	0.47	002.06	0
002	Protein (%)	0049	15.585	0.15000	15.458	0.25981	0.09704	210	0.49	002.06	0
002	Protein (%)	0142	15.585	0.03000	15.458	0.25981	0.09704	210	0.49	002.06	0
002	Protein (%)	2016	15.585	0.21000	15.458	0.25981	0.09704	210	0.49	002.06	0
002	Protein (%)	0958	15.590	0.14000	15.458	0.25981	0.09704	210	0.51	002.06	0
002	Protein (%)	0152	15.600	0.00000	15.458	0.25981	0.09704	210	0.55	002.01	0
002	Protein (%)	0610	15.600	0.00000	15.458	0.25981	0.09704	210	0.55	002.06	0
002	Protein (%)	0956	15.600	0.20000	15.458	0.25981	0.09704	210	0.55	002.06	0
002	Protein (%)	0011	15.610	0.44000	15.458	0.25981	0.09704	210	0.59	002.06	0
002	Protein (%)	0541	15.615	0.29000	15.458	0.25981	0.09704	210	0.61	002.06	0
002	Protein (%)	0006	15.626	0.00400	15.458	0.25981	0.09704	210	0.65	002.06	0
002	Protein (%)	0653	15.630	0.04000	15.458	0.25981	0.09704	210	0.66	002.01	0
002	Protein (%)	0032	15.630	0.06000	15.458	0.25981	0.09704	210	0.66	002.06	0
002	Protein (%)	0512	15.630	0.08000	15.458	0.25981	0.09704	210	0.66	002.06	0
002	Protein (%)	0098	15.650	0.10000	15.458	0.25981	0.09704	210	0.74	002.06	0
002	Protein (%)	0265	15.650	0.10000	15.458	0.25981	0.09704	210	0.74	002.06	0
002	Protein (%)	0294	15.650	0.10000	15.458	0.25981	0.09704	210	0.74	002.06	0
002	Protein (%)	0003	15.650	0.36000	15.458	0.25981	0.09704	210	0.74	002.06	0
002	Protein (%)	0683	15.660	0.00000	15.458	0.25981	0.09704	210	0.78	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
002	Protein (%)	0957	15.680	0.10000	15.458	0.25981	0.09704	210	0.86	002.06	0
002	Protein (%)	2053	15.680	0.06000	15.458	0.25981	0.09704	210	0.86	002.06	0
002	Protein (%)	0018	15.685	0.03000	15.458	0.25981	0.09704	210	0.87	002.06	0
002	Protein (%)	0589	15.685	0.09000	15.458	0.25981	0.09704	210	0.87	002.06	0
002	Protein (%)	0968	15.689	0.11000	15.458	0.25981	0.09704	210	0.89	002.06	0
002	Protein (%)	0106	15.695	0.05000	15.458	0.25981	0.09704	210	0.91	002.00	0
002	Protein (%)	0178	15.700	0.12000	15.458	0.25981	0.09704	210	0.93	002.06	0
002	Protein (%)	0692	15.700	0.00000	15.458	0.25981	0.09704	210	0.93	002.06	0
002	Protein (%)	0233	15.710	0.02000	15.458	0.25981	0.09704	210	0.97	002.06	0
002	Protein (%)	0872	15.725	0.25000	15.458	0.25981	0.09704	210	1.03	002.06	0
002	Protein (%)	0358	15.740	0.38000	15.458	0.25981	0.09704	210	1.09	002.06	0
002	Protein (%)	0001	15.745	0.15000	15.458	0.25981	0.09704	210	1.11	002.06	0
002	Protein (%)	0019	15.745	0.07000	15.458	0.25981	0.09704	210	1.11	002.06	0
002	Protein (%)	0629	15.745	0.01000	15.458	0.25981	0.09704	210	1.11	002.10	0
002	Protein (%)	0610	15.750	0.10000	15.458	0.25981	0.09704	210	1.12	002.01	0
002	Protein (%)	0014	15.750	0.10000	15.458	0.25981	0.09704	210	1.12	002.06	0
002	Protein (%)	0660	15.750	0.04000	15.458	0.25981	0.09704	210	1.12	002.06	0
002	Protein (%)	0169	15.755	0.15000	15.458	0.25981	0.09704	210	1.14	002.02	0
002	Protein (%)	0504	15.765	0.11000	15.458	0.25981	0.09704	210	1.18	002.06	0
002	Protein (%)	0574	15.765	0.01000	15.458	0.25981	0.09704	210	1.18	002.06	0
002	Protein (%)	0941	15.765	0.09000	15.458	0.25981	0.09704	210	1.18	002.06	0
002	Protein (%)	0033	15.775	0.09000	15.458	0.25981	0.09704	210	1.22	002.06	0
002	Protein (%)	0687	15.775	0.05000	15.458	0.25981	0.09704	210	1.22	002.06	0
002	Protein (%)	0843	15.780	0.08000	15.458	0.25981	0.09704	210	1.24	002.06	0
002	Protein (%)	0959	15.780	0.42000	15.458	0.25981	0.09704	210	1.24	002.06	0
002	Protein (%)	0036	15.800	0.04000	15.458	0.25981	0.09704	210	1.32	002.06	0
002	Protein (%)	0964	15.800	0.20000	15.458	0.25981	0.09704	210	1.32	002.06	0
002	Protein (%)	0529	15.805	0.07000	15.458	0.25981	0.09704	210	1.34	002.06	0
002	Protein (%)	0650	15.805	0.01000	15.458	0.25981	0.09704	210	1.34	002.06	0
002	Protein (%)	0425	15.810	0.12000	15.458	0.25981	0.09704	210	1.36	002.06	0
002	Protein (%)	0553	15.820	0.08000	15.458	0.25981	0.09704	210	1.39	002.06	0
002	Protein (%)	0345	15.835	0.09000	15.458	0.25981	0.09704	210	1.45	002.06	0
002	Protein (%)	0034	15.945	0.07700	15.458	0.25981	0.09704	210	1.87	002.06	0
002	Protein (%)	0880	15.945	0.15000	15.458	0.25981	0.09704	210	1.88	002.06	0
002	Protein (%)	0609	16.175	0.09000	15.458	0.25981	0.09704	210	2.76	002.06	0
002	Protein (%)	0961	16.185	0.09000	15.458	0.25981	0.09704	210	2.80	002.06	0
002	Protein (%)	0563	16.302	0.00990	15.458	0.25981	0.09704	210	3.25	002.08	0
002	Protein (%)	0511	16.410	0.02000	15.458	0.25981	0.09704	210	3.67	002.06	0
002	Protein (%)	0646	16.460	0.02000	15.458	0.25981	0.09704	210	3.86	002.06	0
002	Protein (%)	0297	16.595	0.03000	15.458	0.25981	0.09704	210	4.38	002.11	0
002	Protein (%)	0845	16.760	0.00000	15.458	0.25981	0.09704	210	5.01	002.00	0
002	Protein (%)	0918	17.560	0.16000	15.458	0.25981	0.09704	210	8.09	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
002	Protein (%)	0895	17.565	0.01000	15.458	0.25981	0.09704	210	8.11	002.05	0
002	Protein (%)	0520	17.850	0.10000	15.458	0.25981	0.09704	210	9.21	002.06	0
002	Protein (%)	0581	17.935	0.07000	15.458	0.25981	0.09704	210	9.53	002.00	0
002	Protein (%)	0108	15.050	0.50000	15.458	0.25981	0.09704	210	-1.57	002.06	1
002	Protein (%)	0889	15.295	0.69000	15.458	0.25981	0.09704	210	-0.63	002.01	1
002	Protein (%)	0045	15.650	0.50000	15.458	0.25981	0.09704	210	0.74	002.06	1
002	Protein (%)	0508	15.699	0.53330	15.458	0.25981	0.09704	210	0.93	002.06	1
002	Protein (%)	0884	15.900	0.56000	15.458	0.25981	0.09704	210	1.70	002.06	1
002	Protein (%)	0539	2.4200	0.02000	15.458	0.25981	0.09704	210	-50.18	002.06	2
003	Fat (%)	0574	4.4350	0.11000	5.4416	0.22272	0.10248	152	-4.52	003.14	0
003	Fat (%)	0520	4.6300	0.04000	5.4416	0.22272	0.10248	152	-3.64	003.14	0
003	Fat (%)	0357	4.7000	0.20000	5.4416	0.22272	0.10248	152	-3.33	003.14	0
003	Fat (%)	0891	4.7218	0.00110	5.4416	0.22272	0.10248	152	-3.23	003.14	0
003	Fat (%)	0574	4.7350	0.07000	5.4416	0.22272	0.10248	152	-3.17	003.06	0
003	Fat (%)	2068	4.8550	0.07000	5.4416	0.22272	0.10248	152	-2.63	003.09	0
003	Fat (%)	0175	4.8700	0.02000	5.4416	0.22272	0.10248	152	-2.57	003.14	0
003	Fat (%)	2022	4.8800	0.06000	5.4416	0.22272	0.10248	152	-2.52	003.10	0
003	Fat (%)	0004	4.8900	0.08000	5.4416	0.22272	0.10248	152	-2.48	003.09	0
003	Fat (%)	0051	5.0000	0.04000	5.4416	0.22272	0.10248	152	-1.98	003.09	0
003	Fat (%)	0720	5.0050	0.43000	5.4416	0.22272	0.10248	152	-1.96	003.00	0
003	Fat (%)	0609	5.0550	0.21000	5.4416	0.22272	0.10248	152	-1.74	003.10	0
003	Fat (%)	0885	5.1000	0.30000	5.4416	0.22272	0.10248	152	-1.53	003.00	0
003	Fat (%)	0619	5.1000	0.16000	5.4416	0.22272	0.10248	152	-1.53	003.10	0
003	Fat (%)	0208	5.1200	0.08000	5.4416	0.22272	0.10248	152	-1.44	003.13	0
003	Fat (%)	2067	5.1275	0.02100	5.4416	0.22272	0.10248	152	-1.41	003.10	0
003	Fat (%)	0229	5.1300	0.02000	5.4416	0.22272	0.10248	152	-1.40	003.14	0
003	Fat (%)	0529	5.1350	0.01000	5.4416	0.22272	0.10248	152	-1.38	003.14	0
003	Fat (%)	2018	5.1500	0.02000	5.4416	0.22272	0.10248	152	-1.31	003.10	0
003	Fat (%)	0144	5.1650	0.13000	5.4416	0.22272	0.10248	152	-1.24	003.14	0
003	Fat (%)	0861	5.1700	0.02000	5.4416	0.22272	0.10248	152	-1.22	003.10	0
003	Fat (%)	0913	5.1900	0.18000	5.4416	0.22272	0.10248	152	-1.13	003.14	0
003	Fat (%)	0278	5.2000	0.20000	5.4416	0.22272	0.10248	152	-1.08	003.14	0
003	Fat (%)	0009	5.2000	0.10000	5.4416	0.22272	0.10248	152	-1.08	003.14	0
003	Fat (%)	0952	5.2150	0.01000	5.4416	0.22272	0.10248	152	-1.02	003.10	0
003	Fat (%)	0536	5.2200	0.22000	5.4416	0.22272	0.10248	152	-1.00	003.99	0
003	Fat (%)	2062	5.2299	0.04510	5.4416	0.22272	0.10248	152	-0.95	003.12	0
003	Fat (%)	0689	5.2500	0.10000	5.4416	0.22272	0.10248	152	-0.86	003.06	0
003	Fat (%)	0897	5.2550	0.03000	5.4416	0.22272	0.10248	152	-0.84	003.12	0
003	Fat (%)	0598	5.2550	0.01000	5.4416	0.22272	0.10248	152	-0.84	003.14	0
003	Fat (%)	0407	5.2600	0.00000	5.4416	0.22272	0.10248	152	-0.82	003.14	0
003	Fat (%)	0910	5.2700	0.32000	5.4416	0.22272	0.10248	152	-0.77	003.06	0
003	Fat (%)	0629	5.2800	0.04000	5.4416	0.22272	0.10248	152	-0.73	003.10	0



Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
003	Fat (%)	0660	5.2900	0.02000	5.4416	0.22272	0.10248	152	-0.68	003.13	0
003	Fat (%)	0142	5.2950	0.17000	5.4416	0.22272	0.10248	152	-0.66	003.00	0
003	Fat (%)	0202	5.3000	0.02000	5.4416	0.22272	0.10248	152	-0.64	003.14	0
003	Fat (%)	0956	5.3000	0.60000	5.4416	0.22272	0.10248	152	-0.64	003.14	0
003	Fat (%)	0689	5.3000	0.20000	5.4416	0.22272	0.10248	152	-0.64	003.10	0
003	Fat (%)	0868	5.3050	0.05000	5.4416	0.22272	0.10248	152	-0.61	003.10	0
003	Fat (%)	2003	5.3050	0.07000	5.4416	0.22272	0.10248	152	-0.61	003.10	0
003	Fat (%)	0098	5.3200	0.02000	5.4416	0.22272	0.10248	152	-0.55	003.13	0
003	Fat (%)	0337	5.3250	0.01000	5.4416	0.22272	0.10248	152	-0.52	003.00	0
003	Fat (%)	2058	5.3250	0.09000	5.4416	0.22272	0.10248	152	-0.52	003.00	0
003	Fat (%)	2048	5.3300	0.06000	5.4416	0.22272	0.10248	152	-0.50	003.13	0
003	Fat (%)	0011	5.3400	0.00000	5.4416	0.22272	0.10248	152	-0.46	003.06	0
003	Fat (%)	0098	5.3400	0.14000	5.4416	0.22272	0.10248	152	-0.46	003.10	0
003	Fat (%)	0298	5.3400	0.02000	5.4416	0.22272	0.10248	152	-0.46	003.10	0
003	Fat (%)	0693	5.3400	0.02000	5.4416	0.22272	0.10248	152	-0.46	003.10	0
003	Fat (%)	0581	5.3400	0.22000	5.4416	0.22272	0.10248	152	-0.46	003.14	0
003	Fat (%)	0233	5.3500	0.00000	5.4416	0.22272	0.10248	152	-0.41	003.10	0
003	Fat (%)	0265	5.3500	0.10000	5.4416	0.22272	0.10248	152	-0.41	003.14	0
003	Fat (%)	0164	5.3550	0.03000	5.4416	0.22272	0.10248	152	-0.39	003.06	0
003	Fat (%)	0199	5.3550	0.01000	5.4416	0.22272	0.10248	152	-0.39	003.06	0
003	Fat (%)	0033	5.3550	0.01000	5.4416	0.22272	0.10248	152	-0.39	003.13	0
003	Fat (%)	0552	5.3650	0.09000	5.4416	0.22272	0.10248	152	-0.34	003.06	0
003	Fat (%)	0881	5.3680	0.02400	5.4416	0.22272	0.10248	152	-0.33	003.10	0
003	Fat (%)	0425	5.3750	0.01000	5.4416	0.22272	0.10248	152	-0.30	003.06	0
003	Fat (%)	0683	5.3750	0.13000	5.4416	0.22272	0.10248	152	-0.30	003.14	0
003	Fat (%)	2012	5.3800	0.00000	5.4416	0.22272	0.10248	152	-0.28	003.09	0
003	Fat (%)	0948	5.3800	0.62000	5.4416	0.22272	0.10248	152	-0.28	003.14	0
003	Fat (%)	0045	5.3900	0.18000	5.4416	0.22272	0.10248	152	-0.23	003.10	0
003	Fat (%)	0893	5.3950	0.01000	5.4416	0.22272	0.10248	152	-0.21	003.12	0
003	Fat (%)	0015	5.3950	0.09000	5.4416	0.22272	0.10248	152	-0.21	003.14	0
003	Fat (%)	0884	5.4000	0.14000	5.4416	0.22272	0.10248	152	-0.19	003.00	0
003	Fat (%)	0674	5.4000	0.18000	5.4416	0.22272	0.10248	152	-0.19	003.09	0
003	Fat (%)	0178	5.4100	0.14000	5.4416	0.22272	0.10248	152	-0.14	003.10	0
003	Fat (%)	0242	5.4100	0.34000	5.4416	0.22272	0.10248	152	-0.14	003.10	0
003	Fat (%)	0019	5.4150	0.07000	5.4416	0.22272	0.10248	152	-0.12	003.14	0
003	Fat (%)	0878	5.4150	0.05000	5.4416	0.22272	0.10248	152	-0.12	003.14	0
003	Fat (%)	0205	5.4185	0.20300	5.4416	0.22272	0.10248	152	-0.10	003.13	0
003	Fat (%)	0148	5.4200	0.00000	5.4416	0.22272	0.10248	152	-0.10	003.06	0
003	Fat (%)	0685	5.4200	0.04000	5.4416	0.22272	0.10248	152	-0.10	003.09	0
003	Fat (%)	0895	5.4200	0.02000	5.4416	0.22272	0.10248	152	-0.10	003.12	0
003	Fat (%)	0903	5.4250	0.01000	5.4416	0.22272	0.10248	152	-0.07	003.12	0
003	Fat (%)	0294	5.4250	0.03000	5.4416	0.22272	0.10248	152	-0.07	003.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
003	Fat (%)	2065	5.4250	0.00340	5.4416	0.22272	0.10248	152	-0.07	003.12	0
003	Fat (%)	0618	5.4350	0.11000	5.4416	0.22272	0.10248	152	-0.03	003.10	0
003	Fat (%)	0728	5.4350	0.03000	5.4416	0.22272	0.10248	152	-0.03	003.10	0
003	Fat (%)	0590	5.4450	0.21000	5.4416	0.22272	0.10248	152	0.02	003.09	0
003	Fat (%)	0852	5.4500	0.30000	5.4416	0.22272	0.10248	152	0.04	003.06	0
003	Fat (%)	0152	5.4500	0.10000	5.4416	0.22272	0.10248	152	0.04	003.09	0
003	Fat (%)	0089	5.4600	0.00000	5.4416	0.22272	0.10248	152	0.08	003.14	0
003	Fat (%)	0083	5.4650	0.15000	5.4416	0.22272	0.10248	152	0.10	003.06	0
003	Fat (%)	0297	5.4650	0.01000	5.4416	0.22272	0.10248	152	0.10	003.06	0
003	Fat (%)	0187	5.4650	0.01000	5.4416	0.22272	0.10248	152	0.10	003.13	0
003	Fat (%)	0049	5.4700	0.70000	5.4416	0.22272	0.10248	152	0.13	003.14	0
003	Fat (%)	0653	5.4750	0.01000	5.4416	0.22272	0.10248	152	0.15	003.09	0
003	Fat (%)	0646	5.4750	0.07000	5.4416	0.22272	0.10248	152	0.15	003.13	0
003	Fat (%)	0849	5.4815	0.02300	5.4416	0.22272	0.10248	152	0.18	003.06	0
003	Fat (%)	2042	5.4850	0.01000	5.4416	0.22272	0.10248	152	0.19	003.10	0
003	Fat (%)	0171	5.4850	0.01000	5.4416	0.22272	0.10248	152	0.19	003.12	0
003	Fat (%)	0508	5.4855	0.04730	5.4416	0.22272	0.10248	152	0.20	003.09	0
003	Fat (%)	0723	5.4900	0.00000	5.4416	0.22272	0.10248	152	0.22	003.09	0
003	Fat (%)	0169	5.4950	0.27000	5.4416	0.22272	0.10248	152	0.24	003.06	0
003	Fat (%)	2063	5.5100	0.02000	5.4416	0.22272	0.10248	152	0.31	003.06	0
003	Fat (%)	0066	5.5100	0.10000	5.4416	0.22272	0.10248	152	0.31	003.14	0
003	Fat (%)	2023	5.5150	0.27000	5.4416	0.22272	0.10248	152	0.33	003.14	0
003	Fat (%)	0074	5.5200	0.08000	5.4416	0.22272	0.10248	152	0.35	003.06	0
003	Fat (%)	0937	5.5200	0.04000	5.4416	0.22272	0.10248	152	0.35	003.10	0
003	Fat (%)	0300	5.5230	0.40600	5.4416	0.22272	0.10248	152	0.37	003.06	0
003	Fat (%)	0638	5.5250	0.05000	5.4416	0.22272	0.10248	152	0.37	003.09	0
003	Fat (%)	0948	5.5300	0.12000	5.4416	0.22272	0.10248	152	0.40	003.09	0
003	Fat (%)	0108	5.5350	0.07000	5.4416	0.22272	0.10248	152	0.42	003.14	0
003	Fat (%)	0034	5.5400	0.02000	5.4416	0.22272	0.10248	152	0.44	003.10	0
003	Fat (%)	0226	5.5500	0.10000	5.4416	0.22272	0.10248	152	0.49	003.09	0
003	Fat (%)	0676	5.5500	0.10000	5.4416	0.22272	0.10248	152	0.49	003.10	0
003	Fat (%)	0505	5.5600	0.02000	5.4416	0.22272	0.10248	152	0.53	003.09	0
003	Fat (%)	0263	5.5634	0.00670	5.4416	0.22272	0.10248	152	0.55	003.09	0
003	Fat (%)	0675	5.5650	0.03000	5.4416	0.22272	0.10248	152	0.55	003.09	0
003	Fat (%)	0100	5.5700	0.06000	5.4416	0.22272	0.10248	152	0.58	003.10	0
003	Fat (%)	0038	5.5800	0.00000	5.4416	0.22272	0.10248	152	0.62	003.09	0
003	Fat (%)	0354	5.5850	0.03000	5.4416	0.22272	0.10248	152	0.64	003.09	0
003	Fat (%)	0001	5.5900	0.11800	5.4416	0.22272	0.10248	152	0.67	003.14	0
003	Fat (%)	0358	5.5900	0.12000	5.4416	0.22272	0.10248	152	0.67	003.14	0
003	Fat (%)	0047	5.6000	0.00000	5.4416	0.22272	0.10248	152	0.71	003.99	0
003	Fat (%)	0889	5.6050	0.03000	5.4416	0.22272	0.10248	152	0.73	003.10	0
003	Fat (%)	0695	5.6150	0.33000	5.4416	0.22272	0.10248	152	0.78	003.10	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
003	Fat (%)	2011	5.6200	0.04000	5.4416	0.22272	0.10248	152	0.80	003.10	0
003	Fat (%)	0955	5.6350	0.05000	5.4416	0.22272	0.10248	152	0.87	003.14	0
003	Fat (%)	0964	5.6450	0.05000	5.4416	0.22272	0.10248	152	0.91	003.09	0
003	Fat (%)	0510	5.6500	0.10000	5.4416	0.22272	0.10248	152	0.94	003.09	0
003	Fat (%)	0968	5.6525	0.10500	5.4416	0.22272	0.10248	152	0.95	003.09	0
003	Fat (%)	0035	5.6550	0.05000	5.4416	0.22272	0.10248	152	0.96	003.00	0
003	Fat (%)	0939	5.6600	0.18000	5.4416	0.22272	0.10248	152	0.98	003.00	0
003	Fat (%)	0918	5.6600	0.18000	5.4416	0.22272	0.10248	152	0.98	003.06	0
003	Fat (%)	0563	5.6673	0.22370	5.4416	0.22272	0.10248	152	1.01	003.01	0
003	Fat (%)	0504	5.6700	0.22000	5.4416	0.22272	0.10248	152	1.03	003.01	0
003	Fat (%)	0354	5.6900	0.02000	5.4416	0.22272	0.10248	152	1.12	003.00	0
003	Fat (%)	0098	5.6950	0.15000	5.4416	0.22272	0.10248	152	1.14	003.09	0
003	Fat (%)	2006	5.6950	0.03000	5.4416	0.22272	0.10248	152	1.14	003.12	0
003	Fat (%)	0028	5.7000	0.04000	5.4416	0.22272	0.10248	152	1.16	003.13	0
003	Fat (%)	0555	5.7000	0.00000	5.4416	0.22272	0.10248	152	1.16	003.06	0
003	Fat (%)	2053	5.7150	0.03000	5.4416	0.22272	0.10248	152	1.23	003.00	0
003	Fat (%)	0687	5.7750	0.05000	5.4416	0.22272	0.10248	152	1.50	003.06	0
003	Fat (%)	0353	5.7850	0.09000	5.4416	0.22272	0.10248	152	1.54	003.09	0
003	Fat (%)	0175	5.8000	0.04000	5.4416	0.22272	0.10248	152	1.61	003.00	0
003	Fat (%)	0511	5.8050	0.01000	5.4416	0.22272	0.10248	152	1.63	003.06	0
003	Fat (%)	0720	5.8200	0.02000	5.4416	0.22272	0.10248	152	1.70	003.11	0
003	Fat (%)	0139	5.8250	0.15000	5.4416	0.22272	0.10248	152	1.72	003.00	0
003	Fat (%)	0017	5.8500	0.10000	5.4416	0.22272	0.10248	152	1.83	003.00	0
003	Fat (%)	0345	5.8500	0.20000	5.4416	0.22272	0.10248	152	1.83	003.00	0
003	Fat (%)	0726	5.8580	0.00200	5.4416	0.22272	0.10248	152	1.87	003.14	0
003	Fat (%)	0309	5.8808	0.39740	5.4416	0.22272	0.10248	152	1.97	003.00	0
003	Fat (%)	0194	5.9550	0.03000	5.4416	0.22272	0.10248	152	2.30	003.00	0
003	Fat (%)	0413	6.0000	0.40000	5.4416	0.22272	0.10248	152	2.51	003.14	0
003	Fat (%)	0106	6.0800	0.22000	5.4416	0.22272	0.10248	152	2.87	003.01	0
003	Fat (%)	0682	6.0900	0.02000	5.4416	0.22272	0.10248	152	2.91	003.06	0
003	Fat (%)	0848	6.1000	0.40000	5.4416	0.22272	0.10248	152	2.96	003.14	0
003	Fat (%)	0190	6.1300	0.04000	5.4416	0.22272	0.10248	152	3.09	003.14	0
003	Fat (%)	0297	6.1450	0.05000	5.4416	0.22272	0.10248	152	3.16	003.11	0
003	Fat (%)	0065	6.2100	0.02000	5.4416	0.22272	0.10248	152	3.45	003.00	0
003	Fat (%)	0889	6.4100	0.10000	5.4416	0.22272	0.10248	152	4.35	003.11	0
003	Fat (%)	0062	4.5590	0.92000	5.4416	0.22272	0.10248	152	-3.96	003.10	1
003	Fat (%)	0880	6.0500	3.1000	5.4416	0.22272	0.10248	152	2.73	003.99	1
004	Fiber (%)	0720	12.050	0.30000	16.227	1.1868	0.23473	125	-3.52	004.11	0
004	Fiber (%)	0895	13.020	0.02000	16.227	1.1868	0.23473	125	-2.70	004.06	0
004	Fiber (%)	0903	13.105	0.01000	16.227	1.1868	0.23473	125	-2.63	004.06	0
004	Fiber (%)	0265	13.350	0.10000	16.227	1.1868	0.23473	125	-2.42	004.07	0
004	Fiber (%)	0505	13.570	0.12000	16.227	1.1868	0.23473	125	-2.24	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
004	Fiber (%)	0563	13.870	0.05060	16.227	1.1868	0.23473	125	-1.99	004.07	0
004	Fiber (%)	0956	14.000	0.40000	16.227	1.1868	0.23473	125	-1.88	004.07	0
004	Fiber (%)	0098	14.105	0.55000	16.227	1.1868	0.23473	125	-1.79	004.07	0
004	Fiber (%)	0019	14.355	0.25000	16.227	1.1868	0.23473	125	-1.58	004.07	0
004	Fiber (%)	0619	14.400	0.20000	16.227	1.1868	0.23473	125	-1.54	004.07	0
004	Fiber (%)	0596	14.500	0.80000	16.227	1.1868	0.23473	125	-1.46	004.00	0
004	Fiber (%)	0610	14.550	0.10000	16.227	1.1868	0.23473	125	-1.41	004.06	0
004	Fiber (%)	2018	14.550	0.34000	16.227	1.1868	0.23473	125	-1.41	004.07	0
004	Fiber (%)	0178	14.595	0.09000	16.227	1.1868	0.23473	125	-1.38	004.07	0
004	Fiber (%)	0951	14.850	0.14000	16.227	1.1868	0.23473	125	-1.16	004.11	0
004	Fiber (%)	0893	14.955	0.01000	16.227	1.1868	0.23473	125	-1.07	004.06	0
004	Fiber (%)	0194	15.040	0.02000	16.227	1.1868	0.23473	125	-1.00	004.00	0
004	Fiber (%)	0202	15.045	0.03000	16.227	1.1868	0.23473	125	-1.00	004.07	0
004	Fiber (%)	0278	15.050	0.10000	16.227	1.1868	0.23473	125	-0.99	004.07	0
004	Fiber (%)	0553	15.050	0.30000	16.227	1.1868	0.23473	125	-0.99	004.07	0
004	Fiber (%)	0038	15.060	0.08000	16.227	1.1868	0.23473	125	-0.98	004.06	0
004	Fiber (%)	0407	15.070	0.22000	16.227	1.1868	0.23473	125	-0.98	004.07	0
004	Fiber (%)	0520	15.100	0.60000	16.227	1.1868	0.23473	125	-0.95	004.07	0
004	Fiber (%)	0646	15.170	0.52000	16.227	1.1868	0.23473	125	-0.89	004.07	0
004	Fiber (%)	0015	15.195	0.01000	16.227	1.1868	0.23473	125	-0.87	004.07	0
004	Fiber (%)	0581	15.200	0.20000	16.227	1.1868	0.23473	125	-0.87	004.07	0
004	Fiber (%)	0913	15.220	0.22000	16.227	1.1868	0.23473	125	-0.85	004.07	0
004	Fiber (%)	2004	15.250	0.50000	16.227	1.1868	0.23473	125	-0.82	004.00	0
004	Fiber (%)	0683	15.300	0.18000	16.227	1.1868	0.23473	125	-0.78	004.07	0
004	Fiber (%)	0910	15.300	0.00000	16.227	1.1868	0.23473	125	-0.78	004.07	0
004	Fiber (%)	0297	15.315	0.05000	16.227	1.1868	0.23473	125	-0.77	004.07	0
004	Fiber (%)	0918	15.335	0.19000	16.227	1.1868	0.23473	125	-0.75	004.06	0
004	Fiber (%)	0682	15.350	0.10000	16.227	1.1868	0.23473	125	-0.74	004.07	0
004	Fiber (%)	0955	15.350	0.10000	16.227	1.1868	0.23473	125	-0.74	004.07	0
004	Fiber (%)	0708	15.495	0.03000	16.227	1.1868	0.23473	125	-0.62	004.07	0
004	Fiber (%)	0003	15.515	0.19000	16.227	1.1868	0.23473	125	-0.60	004.07	0
004	Fiber (%)	0298	15.590	0.02000	16.227	1.1868	0.23473	125	-0.54	004.00	0
004	Fiber (%)	0552	15.590	0.50000	16.227	1.1868	0.23473	125	-0.54	004.06	0
004	Fiber (%)	0629	15.650	0.10000	16.227	1.1868	0.23473	125	-0.49	004.99	0
004	Fiber (%)	0726	15.684	0.12500	16.227	1.1868	0.23473	125	-0.46	004.07	0
004	Fiber (%)	0171	15.700	0.40000	16.227	1.1868	0.23473	125	-0.44	004.00	0
004	Fiber (%)	0884	15.720	0.48000	16.227	1.1868	0.23473	125	-0.43	004.00	0
004	Fiber (%)	0004	15.750	0.24000	16.227	1.1868	0.23473	125	-0.40	004.07	0
004	Fiber (%)	0510	15.750	0.30000	16.227	1.1868	0.23473	125	-0.40	004.07	0
004	Fiber (%)	0849	15.774	0.18800	16.227	1.1868	0.23473	125	-0.38	004.07	0
004	Fiber (%)	2053	15.790	0.06000	16.227	1.1868	0.23473	125	-0.37	004.07	0
004	Fiber (%)	0098	15.820	0.36000	16.227	1.1868	0.23473	125	-0.34	004.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
004	Fiber (%)	0891	15.844	0.00570	16.227	1.1868	0.23473	125	-0.32	004.07	0
004	Fiber (%)	0685	15.865	0.07000	16.227	1.1868	0.23473	125	-0.31	004.03	0
004	Fiber (%)	0008	15.880	0.48000	16.227	1.1868	0.23473	125	-0.29	004.07	0
004	Fiber (%)	0229	15.900	0.20000	16.227	1.1868	0.23473	125	-0.28	004.07	0
004	Fiber (%)	2022	15.905	0.09000	16.227	1.1868	0.23473	125	-0.27	004.06	0
004	Fiber (%)	0968	16.031	0.20700	16.227	1.1868	0.23473	125	-0.17	004.06	0
004	Fiber (%)	0425	16.050	0.10000	16.227	1.1868	0.23473	125	-0.15	004.00	0
004	Fiber (%)	0897	16.100	0.00000	16.227	1.1868	0.23473	125	-0.11	004.06	0
004	Fiber (%)	0529	16.100	0.04000	16.227	1.1868	0.23473	125	-0.11	004.07	0
004	Fiber (%)	0242	16.145	0.23000	16.227	1.1868	0.23473	125	-0.07	004.07	0
004	Fiber (%)	2012	16.155	0.09000	16.227	1.1868	0.23473	125	-0.06	004.07	0
004	Fiber (%)	0100	16.200	0.10000	16.227	1.1868	0.23473	125	-0.02	004.07	0
004	Fiber (%)	0033	16.200	0.20000	16.227	1.1868	0.23473	125	-0.02	004.07	0
004	Fiber (%)	2058	16.215	0.01000	16.227	1.1868	0.23473	125	-0.01	004.03	0
004	Fiber (%)	0345	16.250	0.50000	16.227	1.1868	0.23473	125	0.02	004.00	0
004	Fiber (%)	0868	16.275	0.01000	16.227	1.1868	0.23473	125	0.04	004.06	0
004	Fiber (%)	0309	16.275	0.02010	16.227	1.1868	0.23473	125	0.04	004.00	0
004	Fiber (%)	0074	16.300	0.02000	16.227	1.1868	0.23473	125	0.06	004.07	0
004	Fiber (%)	0689	16.300	0.60000	16.227	1.1868	0.23473	125	0.06	004.06	0
004	Fiber (%)	0190	16.350	0.10000	16.227	1.1868	0.23473	125	0.10	004.07	0
004	Fiber (%)	0674	16.380	0.06000	16.227	1.1868	0.23473	125	0.13	004.06	0
004	Fiber (%)	0354	16.395	0.01000	16.227	1.1868	0.23473	125	0.14	004.07	0
004	Fiber (%)	0205	16.400	0.08000	16.227	1.1868	0.23473	125	0.15	004.06	0
004	Fiber (%)	0208	16.450	0.50000	16.227	1.1868	0.23473	125	0.19	004.00	0
004	Fiber (%)	0964	16.455	0.73000	16.227	1.1868	0.23473	125	0.19	004.06	0
004	Fiber (%)	0939	16.465	0.33000	16.227	1.1868	0.23473	125	0.20	004.07	0
004	Fiber (%)	0169	16.505	0.33000	16.227	1.1868	0.23473	125	0.23	004.00	0
004	Fiber (%)	2068	16.545	0.09000	16.227	1.1868	0.23473	125	0.27	004.03	0
004	Fiber (%)	0353	16.625	0.21000	16.227	1.1868	0.23473	125	0.34	004.00	0
004	Fiber (%)	2023	16.625	0.39000	16.227	1.1868	0.23473	125	0.34	004.00	0
004	Fiber (%)	0723	16.636	0.00000	16.227	1.1868	0.23473	125	0.34	004.07	0
004	Fiber (%)	0653	16.645	0.01000	16.227	1.1868	0.23473	125	0.35	004.06	0
004	Fiber (%)	0878	16.645	0.11000	16.227	1.1868	0.23473	125	0.35	004.06	0
004	Fiber (%)	0226	16.650	0.10000	16.227	1.1868	0.23473	125	0.36	004.00	0
004	Fiber (%)	2063	16.665	0.03000	16.227	1.1868	0.23473	125	0.37	004.06	0
004	Fiber (%)	0941	16.730	0.14000	16.227	1.1868	0.23473	125	0.42	004.06	0
004	Fiber (%)	0643	16.750	0.44000	16.227	1.1868	0.23473	125	0.44	004.07	0
004	Fiber (%)	0861	16.755	0.27000	16.227	1.1868	0.23473	125	0.44	004.07	0
004	Fiber (%)	2067	16.768	0.02900	16.227	1.1868	0.23473	125	0.46	004.06	0
004	Fiber (%)	0199	16.775	0.07000	16.227	1.1868	0.23473	125	0.46	004.06	0
004	Fiber (%)	0881	16.788	0.20400	16.227	1.1868	0.23473	125	0.47	004.06	0
004	Fiber (%)	0164	16.800	0.20000	16.227	1.1868	0.23473	125	0.48	004.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
004	Fiber (%)	0175	16.800	0.20000	16.227	1.1868	0.23473	125	0.48	004.00	0
004	Fiber (%)	0504	16.865	0.83000	16.227	1.1868	0.23473	125	0.54	004.00	0
004	Fiber (%)	2042	16.875	0.09000	16.227	1.1868	0.23473	125	0.55	004.03	0
004	Fiber (%)	0300	16.899	0.80400	16.227	1.1868	0.23473	125	0.57	004.07	0
004	Fiber (%)	0885	16.905	0.07000	16.227	1.1868	0.23473	125	0.57	004.06	0
004	Fiber (%)	0948	16.910	0.04000	16.227	1.1868	0.23473	125	0.58	004.07	0
004	Fiber (%)	2011	16.965	0.23000	16.227	1.1868	0.23473	125	0.62	004.00	0
004	Fiber (%)	0089	17.020	0.06000	16.227	1.1868	0.23473	125	0.67	004.07	0
004	Fiber (%)	0511	17.150	0.30000	16.227	1.1868	0.23473	125	0.78	004.00	0
004	Fiber (%)	0848	17.205	0.03000	16.227	1.1868	0.23473	125	0.82	004.07	0
004	Fiber (%)	2065	17.253	0.00000	16.227	1.1868	0.23473	125	0.86	004.06	0
004	Fiber (%)	2062	17.259	0.12830	16.227	1.1868	0.23473	125	0.87	004.06	0
004	Fiber (%)	0728	17.300	0.80000	16.227	1.1868	0.23473	125	0.90	004.06	0
004	Fiber (%)	0937	17.305	0.07000	16.227	1.1868	0.23473	125	0.91	004.06	0
004	Fiber (%)	0009	17.390	0.44000	16.227	1.1868	0.23473	125	0.98	004.07	0
004	Fiber (%)	0294	17.400	0.00000	16.227	1.1868	0.23473	125	0.99	004.07	0
004	Fiber (%)	0354	17.420	0.14000	16.227	1.1868	0.23473	125	1.00	004.00	0
004	Fiber (%)	0638	17.450	0.50000	16.227	1.1868	0.23473	125	1.03	004.06	0
004	Fiber (%)	0675	17.495	0.13000	16.227	1.1868	0.23473	125	1.07	004.06	0
004	Fiber (%)	0083	17.535	0.57000	16.227	1.1868	0.23473	125	1.10	004.07	0
004	Fiber (%)	0034	17.550	0.10000	16.227	1.1868	0.23473	125	1.11	004.07	0
004	Fiber (%)	0845	17.640	0.22000	16.227	1.1868	0.23473	125	1.19	004.06	0
004	Fiber (%)	0952	17.690	0.10000	16.227	1.1868	0.23473	125	1.23	004.06	0
004	Fiber (%)	0536	17.755	0.63000	16.227	1.1868	0.23473	125	1.29	004.07	0
004	Fiber (%)	0695	17.840	0.16000	16.227	1.1868	0.23473	125	1.36	004.00	0
004	Fiber (%)	0948	18.190	0.28000	16.227	1.1868	0.23473	125	1.65	004.06	0
004	Fiber (%)	0720	18.255	0.05000	16.227	1.1868	0.23473	125	1.71	004.06	0
004	Fiber (%)	2006	18.490	0.18000	16.227	1.1868	0.23473	125	1.91	004.06	0
004	Fiber (%)	0144	18.550	0.68000	16.227	1.1868	0.23473	125	1.96	004.07	0
004	Fiber (%)	0951	18.560	0.72000	16.227	1.1868	0.23473	125	1.97	004.07	0
004	Fiber (%)	0626	18.565	0.99000	16.227	1.1868	0.23473	125	1.97	004.03	0
004	Fiber (%)	0337	18.690	0.68000	16.227	1.1868	0.23473	125	2.08	004.00	0
004	Fiber (%)	0693	18.875	0.85000	16.227	1.1868	0.23473	125	2.23	004.03	0
004	Fiber (%)	0045	18.900	0.60000	16.227	1.1868	0.23473	125	2.25	004.01	0
004	Fiber (%)	0609	19.115	0.07000	16.227	1.1868	0.23473	125	2.43	004.06	0
004	Fiber (%)	0676	19.750	0.30000	16.227	1.1868	0.23473	125	2.97	004.06	0
004	Fiber (%)	0590	15.630	1.2800	16.227	1.1868	0.23473	125	-0.50	004.06	1
004	Fiber (%)	0413	17.350	1.5000	16.227	1.1868	0.23473	125	0.95	004.07	1
005	Ash (%)	0297	6.0100	0.08000	10.543	0.26869	0.08873	146	-16.87	005.11	0
005	Ash (%)	0889	6.6550	0.19000	10.543	0.26869	0.08873	146	-14.47	005.11	0
005	Ash (%)	0951	6.9850	0.27000	10.543	0.26869	0.08873	146	-13.24	005.11	0
005	Ash (%)	2048	9.2250	0.11000	10.543	0.26869	0.08873	146	-4.90	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
005	Ash (%)	0967	9.2500	0.26000	10.543	0.26869	0.08873	146	-4.81	005.99	0
005	Ash (%)	0720	9.2750	0.07000	10.543	0.26869	0.08873	146	-4.72	005.11	0
005	Ash (%)	0169	9.3300	0.06000	10.543	0.26869	0.08873	146	-4.51	005.00	0
005	Ash (%)	0042	9.3500	0.10000	10.543	0.26869	0.08873	146	-4.44	005.00	0
005	Ash (%)	0650	9.3950	0.23000	10.543	0.26869	0.08873	146	-4.27	005.00	0
005	Ash (%)	0208	9.5050	0.05000	10.543	0.26869	0.08873	146	-3.86	005.00	0
005	Ash (%)	0065	9.6582	0.00630	10.543	0.26869	0.08873	146	-3.29	005.00	0
005	Ash (%)	2067	9.6680	0.00200	10.543	0.26869	0.08873	146	-3.26	005.00	0
005	Ash (%)	0541	9.6850	0.47000	10.543	0.26869	0.08873	146	-3.19	005.00	0
005	Ash (%)	0596	9.7400	0.04000	10.543	0.26869	0.08873	146	-2.99	005.00	0
005	Ash (%)	0539	9.7950	0.21000	10.543	0.26869	0.08873	146	-2.78	005.00	0
005	Ash (%)	0049	9.8300	0.18000	10.543	0.26869	0.08873	146	-2.65	005.00	0
005	Ash (%)	0202	9.8400	0.06000	10.543	0.26869	0.08873	146	-2.62	005.99	0
005	Ash (%)	0959	9.9200	0.22000	10.543	0.26869	0.08873	146	-2.32	005.00	0
005	Ash (%)	0008	9.9750	0.29000	10.543	0.26869	0.08873	146	-2.11	005.00	0
005	Ash (%)	0964	9.9850	0.09000	10.543	0.26869	0.08873	146	-2.08	005.00	0
005	Ash (%)	0598	10.025	0.09000	10.543	0.26869	0.08873	146	-1.93	005.00	0
005	Ash (%)	0194	10.045	0.03000	10.543	0.26869	0.08873	146	-1.85	005.00	0
005	Ash (%)	0142	10.070	0.02000	10.543	0.26869	0.08873	146	-1.76	005.00	0
005	Ash (%)	0893	10.105	0.01000	10.543	0.26869	0.08873	146	-1.63	005.05	0
005	Ash (%)	2068	10.110	0.00000	10.543	0.26869	0.08873	146	-1.61	005.00	0
005	Ash (%)	0001	10.116	0.00750	10.543	0.26869	0.08873	146	-1.59	005.00	0
005	Ash (%)	0205	10.130	0.04000	10.543	0.26869	0.08873	146	-1.54	005.00	0
005	Ash (%)	2058	10.165	0.05000	10.543	0.26869	0.08873	146	-1.41	005.05	0
005	Ash (%)	0015	10.210	0.26000	10.543	0.26869	0.08873	146	-1.24	005.00	0
005	Ash (%)	0874	10.210	0.22000	10.543	0.26869	0.08873	146	-1.24	005.01	0
005	Ash (%)	0178	10.215	0.15000	10.543	0.26869	0.08873	146	-1.22	005.00	0
005	Ash (%)	0918	10.235	0.41000	10.543	0.26869	0.08873	146	-1.15	005.00	0
005	Ash (%)	0861	10.240	0.16000	10.543	0.26869	0.08873	146	-1.13	005.00	0
005	Ash (%)	0175	10.250	0.10000	10.543	0.26869	0.08873	146	-1.09	005.00	0
005	Ash (%)	0957	10.265	0.01000	10.543	0.26869	0.08873	146	-1.03	005.00	0
005	Ash (%)	0953	10.270	0.06000	10.543	0.26869	0.08873	146	-1.01	005.00	0
005	Ash (%)	0609	10.300	0.04000	10.543	0.26869	0.08873	146	-0.90	005.00	0
005	Ash (%)	0520	10.350	0.10000	10.543	0.26869	0.08873	146	-0.72	005.00	0
005	Ash (%)	0881	10.363	0.12300	10.543	0.26869	0.08873	146	-0.67	005.00	0
005	Ash (%)	0425	10.365	0.05000	10.543	0.26869	0.08873	146	-0.66	005.00	0
005	Ash (%)	0100	10.370	0.14000	10.543	0.26869	0.08873	146	-0.64	005.00	0
005	Ash (%)	0958	10.370	0.02000	10.543	0.26869	0.08873	146	-0.64	005.00	0
005	Ash (%)	0968	10.381	0.14600	10.543	0.26869	0.08873	146	-0.60	005.00	0
005	Ash (%)	0038	10.385	0.15000	10.543	0.26869	0.08873	146	-0.59	005.00	0
005	Ash (%)	0278	10.395	0.13000	10.543	0.26869	0.08873	146	-0.55	005.00	0
005	Ash (%)	0171	10.400	0.00000	10.543	0.26869	0.08873	146	-0.53	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
005	Ash (%)	0956	10.400	0.00000	10.543	0.26869	0.08873	146	-0.53	005.00	0
005	Ash (%)	0354	10.405	0.01000	10.543	0.26869	0.08873	146	-0.51	005.00	0
005	Ash (%)	0089	10.445	0.01000	10.543	0.26869	0.08873	146	-0.36	005.00	0
005	Ash (%)	0148	10.455	0.17000	10.543	0.26869	0.08873	146	-0.33	005.00	0
005	Ash (%)	0563	10.458	0.07130	10.543	0.26869	0.08873	146	-0.32	005.00	0
005	Ash (%)	2018	10.460	0.18000	10.543	0.26869	0.08873	146	-0.31	005.05	0
005	Ash (%)	0681	10.485	0.11000	10.543	0.26869	0.08873	146	-0.21	005.00	0
005	Ash (%)	0868	10.485	0.03000	10.543	0.26869	0.08873	146	-0.21	005.05	0
005	Ash (%)	0552	10.495	0.13000	10.543	0.26869	0.08873	146	-0.18	005.00	0
005	Ash (%)	0152	10.500	0.00000	10.543	0.26869	0.08873	146	-0.16	005.00	0
005	Ash (%)	0297	10.500	0.20000	10.543	0.26869	0.08873	146	-0.16	005.00	0
005	Ash (%)	0505	10.515	0.01000	10.543	0.26869	0.08873	146	-0.10	005.00	0
005	Ash (%)	2023	10.515	0.09000	10.543	0.26869	0.08873	146	-0.10	005.99	0
005	Ash (%)	0685	10.520	0.04000	10.543	0.26869	0.08873	146	-0.08	005.05	0
005	Ash (%)	0653	10.530	0.22000	10.543	0.26869	0.08873	146	-0.05	005.00	0
005	Ash (%)	0674	10.540	0.16000	10.543	0.26869	0.08873	146	-0.01	005.00	0
005	Ash (%)	0139	10.550	0.00000	10.543	0.26869	0.08873	146	0.03	005.00	0
005	Ash (%)	0004	10.555	0.03000	10.543	0.26869	0.08873	146	0.05	005.00	0
005	Ash (%)	0051	10.555	0.07000	10.543	0.26869	0.08873	146	0.05	005.00	0
005	Ash (%)	0066	10.555	0.05000	10.543	0.26869	0.08873	146	0.05	005.00	0
005	Ash (%)	0358	10.555	0.21000	10.543	0.26869	0.08873	146	0.05	005.00	0
005	Ash (%)	0643	10.560	0.10000	10.543	0.26869	0.08873	146	0.06	005.00	0
005	Ash (%)	0675	10.560	0.24000	10.543	0.26869	0.08873	146	0.06	005.00	0
005	Ash (%)	2006	10.565	0.05000	10.543	0.26869	0.08873	146	0.08	005.00	0
005	Ash (%)	0536	10.570	0.02000	10.543	0.26869	0.08873	146	0.10	005.99	0
005	Ash (%)	0083	10.585	0.07000	10.543	0.26869	0.08873	146	0.16	005.00	0
005	Ash (%)	0682	10.585	0.17000	10.543	0.26869	0.08873	146	0.16	005.00	0
005	Ash (%)	2012	10.585	0.03000	10.543	0.26869	0.08873	146	0.16	005.00	0
005	Ash (%)	2063	10.590	0.02000	10.543	0.26869	0.08873	146	0.18	005.05	0
005	Ash (%)	0891	10.602	0.17660	10.543	0.26869	0.08873	146	0.22	005.05	0
005	Ash (%)	0618	10.610	0.06000	10.543	0.26869	0.08873	146	0.25	005.00	0
005	Ash (%)	0035	10.620	0.04000	10.543	0.26869	0.08873	146	0.29	005.00	0
005	Ash (%)	0263	10.622	0.04330	10.543	0.26869	0.08873	146	0.29	005.05	0
005	Ash (%)	0353	10.625	0.05000	10.543	0.26869	0.08873	146	0.31	005.00	0
005	Ash (%)	0651	10.627	0.03400	10.543	0.26869	0.08873	146	0.31	005.00	0
005	Ash (%)	0848	10.635	0.01000	10.543	0.26869	0.08873	146	0.34	005.00	0
005	Ash (%)	0581	10.645	0.29000	10.543	0.26869	0.08873	146	0.38	005.00	0
005	Ash (%)	0937	10.645	0.01000	10.543	0.26869	0.08873	146	0.38	005.05	0
005	Ash (%)	2065	10.645	0.00340	10.543	0.26869	0.08873	146	0.38	005.05	0
005	Ash (%)	0676	10.650	0.10000	10.543	0.26869	0.08873	146	0.40	005.99	0
005	Ash (%)	0108	10.660	0.02000	10.543	0.26869	0.08873	146	0.44	005.00	0
005	Ash (%)	0298	10.660	0.02000	10.543	0.26869	0.08873	146	0.44	005.00	0



Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
005	Ash (%)	0728	10.660	0.06000	10.543	0.26869	0.08873	146	0.44	005.99	0
005	Ash (%)	0693	10.670	0.04000	10.543	0.26869	0.08873	146	0.47	005.00	0
005	Ash (%)	0889	10.675	0.11000	10.543	0.26869	0.08873	146	0.49	005.05	0
005	Ash (%)	0574	10.680	0.10000	10.543	0.26869	0.08873	146	0.51	005.00	0
005	Ash (%)	0961	10.680	0.04000	10.543	0.26869	0.08873	146	0.51	005.00	0
005	Ash (%)	0683	10.680	0.30000	10.543	0.26869	0.08873	146	0.51	005.05	0
005	Ash (%)	0723	10.683	0.00000	10.543	0.26869	0.08873	146	0.52	005.00	0
005	Ash (%)	0589	10.690	0.02000	10.543	0.26869	0.08873	146	0.55	005.00	0
005	Ash (%)	0164	10.695	0.07000	10.543	0.26869	0.08873	146	0.57	005.00	0
005	Ash (%)	0660	10.695	0.21000	10.543	0.26869	0.08873	146	0.57	005.00	0
005	Ash (%)	0294	10.700	0.00000	10.543	0.26869	0.08873	146	0.59	005.05	0
005	Ash (%)	0590	10.700	0.08000	10.543	0.26869	0.08873	146	0.59	005.05	0
005	Ash (%)	2042	10.700	0.00000	10.543	0.26869	0.08873	146	0.59	005.05	0
005	Ash (%)	0652	10.700	0.20000	10.543	0.26869	0.08873	146	0.59	005.99	0
005	Ash (%)	2004	10.700	0.00000	10.543	0.26869	0.08873	146	0.59	005.99	0
005	Ash (%)	0913	10.705	0.05000	10.543	0.26869	0.08873	146	0.60	005.00	0
005	Ash (%)	0407	10.710	0.00000	10.543	0.26869	0.08873	146	0.62	005.00	0
005	Ash (%)	0880	10.710	0.30000	10.543	0.26869	0.08873	146	0.62	005.00	0
005	Ash (%)	0265	10.715	0.01000	10.543	0.26869	0.08873	146	0.64	005.00	0
005	Ash (%)	0138	10.716	0.03000	10.543	0.26869	0.08873	146	0.65	005.00	0
005	Ash (%)	2062	10.720	0.14000	10.543	0.26869	0.08873	146	0.66	005.05	0
005	Ash (%)	0638	10.730	0.10000	10.543	0.26869	0.08873	146	0.70	005.05	0
005	Ash (%)	0849	10.738	0.11500	10.543	0.26869	0.08873	146	0.73	005.05	0
005	Ash (%)	0921	10.745	0.07000	10.543	0.26869	0.08873	146	0.75	005.05	0
005	Ash (%)	0629	10.745	0.01000	10.543	0.26869	0.08873	146	0.75	005.99	0
005	Ash (%)	0620	10.748	0.13500	10.543	0.26869	0.08873	146	0.76	005.00	0
005	Ash (%)	0045	10.750	0.10000	10.543	0.26869	0.08873	146	0.77	005.00	0
005	Ash (%)	0510	10.750	0.08000	10.543	0.26869	0.08873	146	0.77	005.00	0
005	Ash (%)	0610	10.750	0.10000	10.543	0.26869	0.08873	146	0.77	005.02	0
005	Ash (%)	0098	10.755	0.01000	10.543	0.26869	0.08873	146	0.79	005.00	0
005	Ash (%)	0646	10.755	0.05000	10.543	0.26869	0.08873	146	0.79	005.00	0
005	Ash (%)	0689	10.755	0.01000	10.543	0.26869	0.08873	146	0.79	005.05	0
005	Ash (%)	0337	10.760	0.04000	10.543	0.26869	0.08873	146	0.81	005.00	0
005	Ash (%)	0033	10.760	0.02000	10.543	0.26869	0.08873	146	0.81	005.05	0
005	Ash (%)	0884	10.760	0.10000	10.543	0.26869	0.08873	146	0.81	005.99	0
005	Ash (%)	2022	10.765	0.09000	10.543	0.26869	0.08873	146	0.83	005.00	0
005	Ash (%)	0948	10.770	0.02000	10.543	0.26869	0.08873	146	0.85	005.00	0
005	Ash (%)	0878	10.770	0.04000	10.543	0.26869	0.08873	146	0.85	005.05	0
005	Ash (%)	0187	10.785	0.03000	10.543	0.26869	0.08873	146	0.90	005.00	0
005	Ash (%)	0903	10.785	0.01000	10.543	0.26869	0.08873	146	0.90	005.05	0
005	Ash (%)	0144	10.800	0.02000	10.543	0.26869	0.08873	146	0.96	005.00	0
005	Ash (%)	0695	10.800	0.04000	10.543	0.26869	0.08873	146	0.96	005.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
005	Ash (%)	0885	10.815	0.03000	10.543	0.26869	0.08873	146	1.01	005.05	0
005	Ash (%)	0345	10.825	0.17000	10.543	0.26869	0.08873	146	1.05	005.00	0
005	Ash (%)	0309	10.838	0.01120	10.543	0.26869	0.08873	146	1.10	005.00	0
005	Ash (%)	0897	10.850	0.10000	10.543	0.26869	0.08873	146	1.14	005.05	0
005	Ash (%)	0242	10.860	0.02000	10.543	0.26869	0.08873	146	1.18	005.00	0
005	Ash (%)	0939	10.865	0.03000	10.543	0.26869	0.08873	146	1.20	005.05	0
005	Ash (%)	0951	10.890	0.18000	10.543	0.26869	0.08873	146	1.29	005.05	0
005	Ash (%)	0357	10.900	0.00000	10.543	0.26869	0.08873	146	1.33	005.00	0
005	Ash (%)	0413	10.900	0.20000	10.543	0.26869	0.08873	146	1.33	005.05	0
005	Ash (%)	0619	10.900	0.00000	10.543	0.26869	0.08873	146	1.33	005.05	0
005	Ash (%)	0229	10.930	0.18000	10.543	0.26869	0.08873	146	1.44	005.00	0
005	Ash (%)	0720	10.940	0.02000	10.543	0.26869	0.08873	146	1.48	005.05	0
005	Ash (%)	0226	10.950	0.10000	10.543	0.26869	0.08873	146	1.52	005.00	0
005	Ash (%)	0895	10.965	0.01000	10.543	0.26869	0.08873	146	1.57	005.05	0
005	Ash (%)	2011	11.040	0.10000	10.543	0.26869	0.08873	146	1.85	005.05	0
005	Ash (%)	0852	11.300	0.00000	10.543	0.26869	0.08873	146	2.82	005.00	0
005	Ash (%)	0960	9.7650	0.75000	10.543	0.26869	0.08873	146	-2.89	005.00	1
005	Ash (%)	0623	9.9701	0.57540	10.543	0.26869	0.08873	146	-2.13	005.00	1
005	Ash (%)	0553	10.285	0.81000	10.543	0.26869	0.08873	146	-0.96	005.00	1
005	Ash (%)	0504	10.870	0.54000	10.543	0.26869	0.08873	146	1.22	005.00	1
005	Ash (%)	0062	11.617	0.67300	10.543	0.26869	0.08873	146	4.00	005.00	1
005	Ash (%)	0845	5.0700	0.08000	10.543	0.26869	0.08873	146	-20.37	005.00	2
006	Total sugars (%)	0910	0.00000	0.00000	4.1338	1.3486	0.18429	14	-3.07	006.00	0
006	Total sugars (%)	0638	2.3000	0.00000	4.1338	1.3486	0.18429	14	-1.36	006.99	0
006	Total sugars (%)	0227	2.3800	0.22000	4.1338	1.3486	0.18429	14	-1.30	006.99	0
006	Total sugars (%)	0921	2.4850	0.65000	4.1338	1.3486	0.18429	14	-1.22	006.00	0
006	Total sugars (%)	0723	3.5600	0.00000	4.1338	1.3486	0.18429	14	-0.43	006.99	0
006	Total sugars (%)	0939	3.6050	0.21000	4.1338	1.3486	0.18429	14	-0.39	006.00	0
006	Total sugars (%)	0689	3.9000	0.20000	4.1338	1.3486	0.18429	14	-0.17	006.01	0
006	Total sugars (%)	0956	4.7000	0.40000	4.1338	1.3486	0.18429	14	0.42	006.99	0
006	Total sugars (%)	0720	4.9100	0.02000	4.1338	1.3486	0.18429	14	0.58	006.99	0
006	Total sugars (%)	0407	4.9650	0.15000	4.1338	1.3486	0.18429	14	0.62	006.01	0
006	Total sugars (%)	0066	5.0100	0.60000	4.1338	1.3486	0.18429	14	0.65	006.00	0
006	Total sugars (%)	0226	5.0500	0.10000	4.1338	1.3486	0.18429	14	0.68	006.99	0
006	Total sugars (%)	0951	5.6450	0.03000	4.1338	1.3486	0.18429	14	1.12	006.99	0
006	Total sugars (%)	0510	6.3000	0.00000	4.1338	1.3486	0.18429	14	1.61	006.99	0
008	Fiber (%)	0354	17.315	0.03000	21.075	2.1083	0.33161	44	-1.78	008.08	0
008	Fiber (%)	0049	18.305	0.09000	21.075	2.1083	0.33161	44	-1.31	008.08	0
008	Fiber (%)	0178	18.570	0.80000	21.075	2.1083	0.33161	44	-1.19	008.08	0
008	Fiber (%)	2004	18.700	0.40000	21.075	2.1083	0.33161	44	-1.13	008.08	0
008	Fiber (%)	0878	18.750	0.64000	21.075	2.1083	0.33161	44	-1.10	008.08	0
008	Fiber (%)	0918	18.775	0.49000	21.075	2.1083	0.33161	44	-1.09	008.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
008	Fiber (%)	0861	18.815	0.43000	21.075	2.1083	0.33161	44	-1.07	008.08	0
008	Fiber (%)	0278	18.850	0.10000	21.075	2.1083	0.33161	44	-1.06	008.08	0
008	Fiber (%)	0164	19.350	0.30000	21.075	2.1083	0.33161	44	-0.82	008.08	0
008	Fiber (%)	0720	19.375	0.13000	21.075	2.1083	0.33161	44	-0.81	008.99	0
008	Fiber (%)	0106	19.475	0.07000	21.075	2.1083	0.33161	44	-0.76	008.08	0
008	Fiber (%)	0263	19.662	0.11670	21.075	2.1083	0.33161	44	-0.67	008.08	0
008	Fiber (%)	0407	19.715	0.15000	21.075	2.1083	0.33161	44	-0.64	008.08	0
008	Fiber (%)	0723	19.800	0.00000	21.075	2.1083	0.33161	44	-0.60	008.08	0
008	Fiber (%)	0357	19.850	0.10000	21.075	2.1083	0.33161	44	-0.58	008.08	0
008	Fiber (%)	0581	19.890	0.50000	21.075	2.1083	0.33161	44	-0.56	008.08	0
008	Fiber (%)	0001	20.075	0.21000	21.075	2.1083	0.33161	44	-0.47	008.08	0
008	Fiber (%)	0148	20.215	0.25000	21.075	2.1083	0.33161	44	-0.41	008.08	0
008	Fiber (%)	0504	20.455	0.87000	21.075	2.1083	0.33161	44	-0.29	008.02	0
008	Fiber (%)	0948	20.495	0.01000	21.075	2.1083	0.33161	44	-0.27	008.02	0
008	Fiber (%)	2012	20.605	0.07000	21.075	2.1083	0.33161	44	-0.22	008.08	0
008	Fiber (%)	0358	20.730	0.34000	21.075	2.1083	0.33161	44	-0.16	008.08	0
008	Fiber (%)	0646	20.730	0.04000	21.075	2.1083	0.33161	44	-0.16	008.08	0
008	Fiber (%)	0413	21.350	1.3000	21.075	2.1083	0.33161	44	0.13	008.08	0
008	Fiber (%)	0066	21.445	0.57000	21.075	2.1083	0.33161	44	0.18	008.08	0
008	Fiber (%)	0510	21.600	0.00000	21.075	2.1083	0.33161	44	0.25	008.08	0
008	Fiber (%)	0265	21.650	0.30000	21.075	2.1083	0.33161	44	0.27	008.05	0
008	Fiber (%)	0297	21.890	0.06000	21.075	2.1083	0.33161	44	0.39	008.08	0
008	Fiber (%)	0353	22.220	0.04000	21.075	2.1083	0.33161	44	0.54	008.02	0
008	Fiber (%)	0098	22.235	1.1700	21.075	2.1083	0.33161	44	0.55	008.02	0
008	Fiber (%)	0309	22.270	0.29400	21.075	2.1083	0.33161	44	0.57	008.02	0
008	Fiber (%)	0038	22.300	0.10000	21.075	2.1083	0.33161	44	0.58	008.02	0
008	Fiber (%)	0083	22.550	1.1000	21.075	2.1083	0.33161	44	0.70	008.08	0
008	Fiber (%)	0675	22.760	0.06000	21.075	2.1083	0.33161	44	0.80	008.02	0
008	Fiber (%)	0148	22.785	0.01000	21.075	2.1083	0.33161	44	0.81	008.02	0
008	Fiber (%)	0045	22.900	1.0000	21.075	2.1083	0.33161	44	0.87	008.08	0
008	Fiber (%)	0294	23.000	0.00000	21.075	2.1083	0.33161	44	0.91	008.08	0
008	Fiber (%)	0037	23.080	0.06000	21.075	2.1083	0.33161	44	0.95	008.08	0
008	Fiber (%)	0693	23.570	0.80000	21.075	2.1083	0.33161	44	1.18	008.08	0
008	Fiber (%)	0226	24.250	0.30000	21.075	2.1083	0.33161	44	1.51	008.02	0
008	Fiber (%)	0187	24.415	0.03000	21.075	2.1083	0.33161	44	1.58	008.02	0
008	Fiber (%)	0609	25.455	0.09000	21.075	2.1083	0.33161	44	2.08	008.02	0
008	Fiber (%)	0405	31.815	0.03000	21.075	2.1083	0.33161	44	5.09	008.02	0
008	Fiber (%)	0536	34.400	1.1400	21.075	2.1083	0.33161	44	6.32	008.08	0
008	Fiber (%)	0968	16.315	2.1100	21.075	2.1083	0.33161	44	-2.26	008.08	1
008	Fiber (%)	0956	18.800	2.0000	21.075	2.1083	0.33161	44	-1.08	008.08	1
009	Fiber (%)	0956	30.150	0.70000	34.999	2.9282	0.52283	40	-1.66	009.09	0
009	Fiber (%)	0278	30.850	0.70000	34.999	2.9282	0.52283	40	-1.42	009.09	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
009	Fiber (%)	0354	31.070	0.04000	34.999	2.9282	0.52283	40	-1.34	009.09	0
009	Fiber (%)	0646	31.435	0.23000	34.999	2.9282	0.52283	40	-1.22	009.09	0
009	Fiber (%)	0861	31.440	0.30000	34.999	2.9282	0.52283	40	-1.22	009.09	0
009	Fiber (%)	2004	31.550	0.90000	34.999	2.9282	0.52283	40	-1.18	009.09	0
009	Fiber (%)	0265	31.600	0.40000	34.999	2.9282	0.52283	40	-1.16	009.09	0
009	Fiber (%)	0878	32.025	0.31000	34.999	2.9282	0.52283	40	-1.02	009.09	0
009	Fiber (%)	0106	32.675	0.51000	34.999	2.9282	0.52283	40	-0.79	009.09	0
009	Fiber (%)	0164	32.850	0.10000	34.999	2.9282	0.52283	40	-0.73	009.09	0
009	Fiber (%)	0504	33.010	0.56000	34.999	2.9282	0.52283	40	-0.68	009.04	0
009	Fiber (%)	0357	33.050	1.3000	34.999	2.9282	0.52283	40	-0.67	009.09	0
009	Fiber (%)	0407	33.185	1.1700	34.999	2.9282	0.52283	40	-0.62	009.09	0
009	Fiber (%)	0098	33.350	2.5000	34.999	2.9282	0.52283	40	-0.56	009.07	0
009	Fiber (%)	0001	33.430	0.58000	34.999	2.9282	0.52283	40	-0.54	009.09	0
009	Fiber (%)	0581	33.910	0.78000	34.999	2.9282	0.52283	40	-0.37	009.09	0
009	Fiber (%)	0510	34.150	0.30000	34.999	2.9282	0.52283	40	-0.29	009.09	0
009	Fiber (%)	0413	34.700	2.6000	34.999	2.9282	0.52283	40	-0.10	009.09	0
009	Fiber (%)	0723	35.250	0.00000	34.999	2.9282	0.52283	40	0.09	009.09	0
009	Fiber (%)	0038	35.345	0.17000	34.999	2.9282	0.52283	40	0.12	009.07	0
009	Fiber (%)	0918	35.460	0.36000	34.999	2.9282	0.52283	40	0.16	009.09	0
009	Fiber (%)	2012	35.715	0.41000	34.999	2.9282	0.52283	40	0.24	009.09	0
009	Fiber (%)	0353	35.950	0.06000	34.999	2.9282	0.52283	40	0.32	009.07	0
009	Fiber (%)	0720	36.070	0.04000	34.999	2.9282	0.52283	40	0.37	009.99	0
009	Fiber (%)	0948	36.220	0.06000	34.999	2.9282	0.52283	40	0.42	009.09	0
009	Fiber (%)	0178	36.430	0.60000	34.999	2.9282	0.52283	40	0.49	009.09	0
009	Fiber (%)	0187	36.705	0.07000	34.999	2.9282	0.52283	40	0.58	009.07	0
009	Fiber (%)	0297	36.770	0.30000	34.999	2.9282	0.52283	40	0.60	009.09	0
009	Fiber (%)	0037	36.800	0.32000	34.999	2.9282	0.52283	40	0.61	009.09	0
009	Fiber (%)	0083	36.900	1.2000	34.999	2.9282	0.52283	40	0.65	009.09	0
009	Fiber (%)	0693	36.910	0.74000	34.999	2.9282	0.52283	40	0.65	009.09	0
009	Fiber (%)	0045	37.050	0.10000	34.999	2.9282	0.52283	40	0.70	009.09	0
009	Fiber (%)	0226	37.100	0.20000	34.999	2.9282	0.52283	40	0.72	009.07	0
009	Fiber (%)	0294	37.300	0.00000	34.999	2.9282	0.52283	40	0.79	009.09	0
009	Fiber (%)	0309	37.418	0.34300	34.999	2.9282	0.52283	40	0.83	009.07	0
009	Fiber (%)	0889	37.930	0.30000	34.999	2.9282	0.52283	40	1.00	009.09	0
009	Fiber (%)	0675	39.110	0.08000	34.999	2.9282	0.52283	40	1.40	009.07	0
009	Fiber (%)	0066	39.655	1.0100	34.999	2.9282	0.52283	40	1.59	009.09	0
009	Fiber (%)	0609	41.200	0.12000	34.999	2.9282	0.52283	40	2.12	009.07	0
009	Fiber (%)	0536	44.855	0.45000	34.999	2.9282	0.52283	40	3.37	009.09	0
009	Fiber (%)	0968	31.220	4.5400	34.999	2.9282	0.52283	40	-1.29	009.09	1
009	Fiber (%)	0049	33.010	6.5600	34.999	2.9282	0.52283	40	-0.68	009.09	1
010	Moisture (%)	0967	6.2500	0.04000	7.1303	0.61578	0.09111	18	-1.43	010.99	0
010	Moisture (%)	0168	6.2800	0.14000	7.1303	0.61578	0.09111	18	-1.38	010.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
010	Moisture (%)	0964	6.6150	0.23000	7.1303	0.61578	0.09111	18	-0.84	010.99	0
010	Moisture (%)	0164	6.7050	0.27000	7.1303	0.61578	0.09111	18	-0.69	010.03	0
010	Moisture (%)	0843	6.7150	0.17000	7.1303	0.61578	0.09111	18	-0.67	010.03	0
010	Moisture (%)	0208	6.7850	0.05000	7.1303	0.61578	0.09111	18	-0.56	010.03	0
010	Moisture (%)	0948	6.9950	0.01000	7.1303	0.61578	0.09111	18	-0.22	010.99	0
010	Moisture (%)	0952	7.0100	0.08000	7.1303	0.61578	0.09111	18	-0.20	010.99	0
010	Moisture (%)	0939	7.0550	0.03000	7.1303	0.61578	0.09111	18	-0.12	010.99	0
010	Moisture (%)	2043	7.1250	0.13000	7.1303	0.61578	0.09111	18	-0.01	010.11	0
010	Moisture (%)	0884	7.1700	0.02000	7.1303	0.61578	0.09111	18	0.06	010.99	0
010	Moisture (%)	2004	7.1800	0.02000	7.1303	0.61578	0.09111	18	0.08	010.99	0
010	Moisture (%)	0889	7.2250	0.05000	7.1303	0.61578	0.09111	18	0.15	010.11	0
010	Moisture (%)	0921	7.4950	0.07000	7.1303	0.61578	0.09111	18	0.59	010.99	0
010	Moisture (%)	0405	8.0850	0.03000	7.1303	0.61578	0.09111	18	1.55	010.99	0
010	Moisture (%)	0720	8.2100	0.06000	7.1303	0.61578	0.09111	18	1.75	010.11	0
010	Moisture (%)	0951	8.3750	0.01000	7.1303	0.61578	0.09111	18	2.02	010.11	0
010	Moisture (%)	0553	8.4350	0.23000	7.1303	0.61578	0.09111	18	2.12	010.99	0
011	Loss on Drying (%)	0874	6.8950	0.35000	8.1113	0.39383	0.12210	76	-3.09	011.01	0
011	Loss on Drying (%)	0646	6.9400	0.06000	8.1113	0.39383	0.12210	76	-2.97	011.01	0
011	Loss on Drying (%)	0675	7.0800	0.18000	8.1113	0.39383	0.12210	76	-2.62	011.01	0
011	Loss on Drying (%)	2016	7.1500	0.26000	8.1113	0.39383	0.12210	76	-2.44	011.01	0
011	Loss on Drying (%)	0685	7.2700	0.00000	8.1113	0.39383	0.12210	76	-2.14	011.01	0
011	Loss on Drying (%)	0541	7.4250	0.15000	8.1113	0.39383	0.12210	76	-1.74	011.99	0
011	Loss on Drying (%)	0953	7.4850	0.05000	8.1113	0.39383	0.12210	76	-1.59	011.01	0
011	Loss on Drying (%)	0175	7.5000	0.20000	8.1113	0.39383	0.12210	76	-1.55	011.01	0
011	Loss on Drying (%)	0178	7.5750	0.51000	8.1113	0.39383	0.12210	76	-1.36	011.01	0
011	Loss on Drying (%)	0942	7.5900	0.02000	8.1113	0.39383	0.12210	76	-1.32	011.02	0
011	Loss on Drying (%)	0843	7.6450	0.05000	8.1113	0.39383	0.12210	76	-1.18	011.01	0
011	Loss on Drying (%)	0051	7.6850	0.13000	8.1113	0.39383	0.12210	76	-1.08	011.01	0
011	Loss on Drying (%)	0881	7.7185	0.11900	8.1113	0.39383	0.12210	76	-1.00	011.01	0
011	Loss on Drying (%)	0294	7.7500	0.10000	8.1113	0.39383	0.12210	76	-0.92	011.01	0
011	Loss on Drying (%)	0598	7.7800	0.02000	8.1113	0.39383	0.12210	76	-0.84	011.01	0
011	Loss on Drying (%)	0848	7.7800	0.26000	8.1113	0.39383	0.12210	76	-0.84	011.01	0
011	Loss on Drying (%)	0062	7.8420	0.28200	8.1113	0.39383	0.12210	76	-0.68	011.01	0
011	Loss on Drying (%)	0968	7.8465	0.04700	8.1113	0.39383	0.12210	76	-0.67	011.01	0
011	Loss on Drying (%)	0171	7.8650	0.11000	8.1113	0.39383	0.12210	76	-0.63	011.01	0
011	Loss on Drying (%)	0574	7.8700	0.32000	8.1113	0.39383	0.12210	76	-0.61	011.01	0
011	Loss on Drying (%)	0066	7.8850	0.07000	8.1113	0.39383	0.12210	76	-0.57	011.01	0
011	Loss on Drying (%)	0653	7.8900	0.00000	8.1113	0.39383	0.12210	76	-0.56	011.01	0
011	Loss on Drying (%)	0857	7.9000	0.20000	8.1113	0.39383	0.12210	76	-0.54	011.99	0
011	Loss on Drying (%)	0354	7.9450	0.01000	8.1113	0.39383	0.12210	76	-0.42	011.01	0
011	Loss on Drying (%)	0510	7.9500	0.10000	8.1113	0.39383	0.12210	76	-0.41	011.01	0
011	Loss on Drying (%)	0651	7.9740	0.00200	8.1113	0.39383	0.12210	76	-0.35	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
011	Loss on Drying (%)	0596	7.9750	0.21000	8.1113	0.39383	0.12210	76	-0.35	011.01	0
011	Loss on Drying (%)	0620	7.9809	0.00660	8.1113	0.39383	0.12210	76	-0.33	011.01	0
011	Loss on Drying (%)	0298	7.9900	0.02000	8.1113	0.39383	0.12210	76	-0.31	011.01	0
011	Loss on Drying (%)	0682	8.0100	0.04000	8.1113	0.39383	0.12210	76	-0.26	011.01	0
011	Loss on Drying (%)	0880	8.0200	0.12000	8.1113	0.39383	0.12210	76	-0.23	011.01	0
011	Loss on Drying (%)	0650	8.0350	0.01000	8.1113	0.39383	0.12210	76	-0.19	011.01	0
011	Loss on Drying (%)	0728	8.0450	0.17000	8.1113	0.39383	0.12210	76	-0.17	011.01	0
011	Loss on Drying (%)	0226	8.0500	0.10000	8.1113	0.39383	0.12210	76	-0.16	011.01	0
011	Loss on Drying (%)	0520	8.0500	0.10000	8.1113	0.39383	0.12210	76	-0.16	011.01	0
011	Loss on Drying (%)	0681	8.0550	0.05000	8.1113	0.39383	0.12210	76	-0.14	011.03	0
011	Loss on Drying (%)	0108	8.1000	0.36000	8.1113	0.39383	0.12210	76	-0.03	011.01	0
011	Loss on Drying (%)	0242	8.1050	0.21000	8.1113	0.39383	0.12210	76	-0.02	011.01	0
011	Loss on Drying (%)	2006	8.1300	0.04000	8.1113	0.39383	0.12210	76	0.05	011.01	0
011	Loss on Drying (%)	0852	8.1400	0.06000	8.1113	0.39383	0.12210	76	0.07	011.99	0
011	Loss on Drying (%)	0152	8.1500	0.10000	8.1113	0.39383	0.12210	76	0.10	011.01	0
011	Loss on Drying (%)	0674	8.1550	0.21000	8.1113	0.39383	0.12210	76	0.11	011.01	0
011	Loss on Drying (%)	0539	8.1700	0.22000	8.1113	0.39383	0.12210	76	0.15	011.01	0
011	Loss on Drying (%)	0033	8.1850	0.17000	8.1113	0.39383	0.12210	76	0.19	011.01	0
011	Loss on Drying (%)	0208	8.1850	0.23000	8.1113	0.39383	0.12210	76	0.19	011.01	0
011	Loss on Drying (%)	2023	8.1850	0.05000	8.1113	0.39383	0.12210	76	0.19	011.02	0
011	Loss on Drying (%)	0265	8.2000	0.20000	8.1113	0.39383	0.12210	76	0.23	011.01	0
011	Loss on Drying (%)	0358	8.2000	0.12000	8.1113	0.39383	0.12210	76	0.23	011.01	0
011	Loss on Drying (%)	0552	8.2150	0.01000	8.1113	0.39383	0.12210	76	0.26	011.01	0
011	Loss on Drying (%)	0194	8.2250	0.03000	8.1113	0.39383	0.12210	76	0.29	011.01	0
011	Loss on Drying (%)	0008	8.2350	0.05000	8.1113	0.39383	0.12210	76	0.31	011.01	0
011	Loss on Drying (%)	0144	8.2350	0.13000	8.1113	0.39383	0.12210	76	0.31	011.01	0
011	Loss on Drying (%)	0148	8.2550	0.03000	8.1113	0.39383	0.12210	76	0.36	011.01	0
011	Loss on Drying (%)	0957	8.2600	0.06000	8.1113	0.39383	0.12210	76	0.38	011.01	0
011	Loss on Drying (%)	0138	8.2830	0.03600	8.1113	0.39383	0.12210	76	0.44	011.01	0
011	Loss on Drying (%)	0589	8.3350	0.29000	8.1113	0.39383	0.12210	76	0.57	011.01	0
011	Loss on Drying (%)	0098	8.3450	0.23000	8.1113	0.39383	0.12210	76	0.59	011.01	0
011	Loss on Drying (%)	0961	8.3950	0.05000	8.1113	0.39383	0.12210	76	0.72	011.01	0
011	Loss on Drying (%)	0164	8.4050	0.09000	8.1113	0.39383	0.12210	76	0.75	011.01	0
011	Loss on Drying (%)	0233	8.4300	0.02000	8.1113	0.39383	0.12210	76	0.81	011.01	0
011	Loss on Drying (%)	0660	8.4450	0.17000	8.1113	0.39383	0.12210	76	0.85	011.01	0
011	Loss on Drying (%)	0511	8.4650	0.11000	8.1113	0.39383	0.12210	76	0.90	011.01	0
011	Loss on Drying (%)	0100	8.4850	0.11000	8.1113	0.39383	0.12210	76	0.95	011.01	0
011	Loss on Drying (%)	0861	8.5650	0.05000	8.1113	0.39383	0.12210	76	1.15	011.01	0
011	Loss on Drying (%)	0723	8.6000	0.00000	8.1113	0.39383	0.12210	76	1.24	011.01	0
011	Loss on Drying (%)	0202	8.6000	0.10000	8.1113	0.39383	0.12210	76	1.24	011.01	0
011	Loss on Drying (%)	0205	8.6045	0.10500	8.1113	0.39383	0.12210	76	1.25	011.01	0
011	Loss on Drying (%)	0229	8.6100	0.02000	8.1113	0.39383	0.12210	76	1.27	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
011	Loss on Drying (%)	0309	8.6500	0.16000	8.1113	0.39383	0.12210	76	1.37	011.01	0
011	Loss on Drying (%)	0958	8.6900	0.14000	8.1113	0.39383	0.12210	76	1.47	011.01	0
011	Loss on Drying (%)	0541	8.6900	0.42000	8.1113	0.39383	0.12210	76	1.47	011.01	0
011	Loss on Drying (%)	0959	8.7200	0.08000	8.1113	0.39383	0.12210	76	1.55	011.01	0
011	Loss on Drying (%)	2067	8.7320	0.00200	8.1113	0.39383	0.12210	76	1.58	011.01	0
011	Loss on Drying (%)	0913	8.8100	0.18000	8.1113	0.39383	0.12210	76	1.77	011.01	0
011	Loss on Drying (%)	2022	8.8950	0.13000	8.1113	0.39383	0.12210	76	1.99	011.01	0
011	Loss on Drying (%)	0407	9.0400	0.08000	8.1113	0.39383	0.12210	76	2.36	011.01	0
011	Loss on Drying (%)	0960	7.6850	0.75000	8.1113	0.39383	0.12210	76	-1.08	011.01	1
012	Starch (%)	0510	10.600	0.60000	13.179	0.87393	0.16935	31	-2.95	012.04	0
012	Starch (%)	0918	11.272	0.50000	13.179	0.87393	0.16935	31	-2.18	012.01	0
012	Starch (%)	0885	11.375	0.01000	13.179	0.87393	0.16935	31	-2.06	012.03	0
012	Starch (%)	0407	11.505	0.71000	13.179	0.87393	0.16935	31	-1.92	012.03	0
012	Starch (%)	0676	11.850	0.10000	13.179	0.87393	0.16935	31	-1.52	012.01	0
012	Starch (%)	0682	12.270	0.38000	13.179	0.87393	0.16935	31	-1.04	012.00	0
012	Starch (%)	0956	12.550	0.30000	13.179	0.87393	0.16935	31	-0.72	012.01	0
012	Starch (%)	0297	12.640	0.02000	13.179	0.87393	0.16935	31	-0.62	012.03	0
012	Starch (%)	0265	12.650	0.10000	13.179	0.87393	0.16935	31	-0.61	012.01	0
012	Starch (%)	0278	12.800	0.20000	13.179	0.87393	0.16935	31	-0.43	012.04	0
012	Starch (%)	0098	12.900	0.00000	13.179	0.87393	0.16935	31	-0.32	012.03	0
012	Starch (%)	0951	13.040	0.22000	13.179	0.87393	0.16935	31	-0.16	012.00	0
012	Starch (%)	0353	13.050	0.08000	13.179	0.87393	0.16935	31	-0.15	012.04	0
012	Starch (%)	0178	13.150	0.22000	13.179	0.87393	0.16935	31	-0.03	012.00	0
012	Starch (%)	0609	13.170	0.00000	13.179	0.87393	0.16935	31	-0.01	012.00	0
012	Starch (%)	0297	13.235	0.11000	13.179	0.87393	0.16935	31	0.06	012.11	0
012	Starch (%)	0638	13.300	0.00000	13.179	0.87393	0.16935	31	0.14	012.00	0
012	Starch (%)	0354	13.360	0.02000	13.179	0.87393	0.16935	31	0.21	012.00	0
012	Starch (%)	0889	13.385	0.09000	13.179	0.87393	0.16935	31	0.24	012.11	0
012	Starch (%)	2023	13.450	0.10000	13.179	0.87393	0.16935	31	0.31	012.00	0
012	Starch (%)	0683	13.505	0.17000	13.179	0.87393	0.16935	31	0.37	012.00	0
012	Starch (%)	0920	13.510	0.04000	13.179	0.87393	0.16935	31	0.38	012.00	0
012	Starch (%)	0723	13.637	0.00000	13.179	0.87393	0.16935	31	0.52	012.00	0
012	Starch (%)	0106	13.715	0.47000	13.179	0.87393	0.16935	31	0.61	012.04	0
012	Starch (%)	0878	13.800	0.20000	13.179	0.87393	0.16935	31	0.71	012.00	0
012	Starch (%)	0619	13.850	0.10000	13.179	0.87393	0.16935	31	0.77	012.00	0
012	Starch (%)	0948	14.185	0.05000	13.179	0.87393	0.16935	31	1.15	012.04	0
012	Starch (%)	0689	14.200	0.00000	13.179	0.87393	0.16935	31	1.17	012.00	0
012	Starch (%)	0720	14.340	0.02000	13.179	0.87393	0.16935	31	1.33	012.11	0
012	Starch (%)	0951	15.500	0.04000	13.179	0.87393	0.16935	31	2.66	012.11	0
012	Starch (%)	2004	16.800	0.40000	13.179	0.87393	0.16935	31	4.14	012.01	0
013	Fat (%)	0845	3.8300	0.24000	6.4767	0.49843	0.19555	53	-5.31	013.10	0
013	Fat (%)	0618	4.7050	0.23000	6.4767	0.49843	0.19555	53	-3.55	013.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
013 Fat (%)		0951	5.3750	0.11000	6.4767	0.49843	0.19555	53	-2.21	013.00	0
013 Fat (%)		0033	5.5600	0.18000	6.4767	0.49843	0.19555	53	-1.84	013.10	0
013 Fat (%)		0874	5.7950	0.13000	6.4767	0.49843	0.19555	53	-1.37	013.00	0
013 Fat (%)		2003	5.8650	0.07000	6.4767	0.49843	0.19555	53	-1.23	013.10	0
013 Fat (%)		0083	5.9550	0.13000	6.4767	0.49843	0.19555	53	-1.05	013.00	0
013 Fat (%)		0958	5.9900	0.06000	6.4767	0.49843	0.19555	53	-0.98	013.02	0
013 Fat (%)		0884	6.0050	0.15000	6.4767	0.49843	0.19555	53	-0.95	013.00	0
013 Fat (%)		0148	6.0050	0.01000	6.4767	0.49843	0.19555	53	-0.95	013.02	0
013 Fat (%)		0610	6.0700	0.20000	6.4767	0.49843	0.19555	53	-0.82	013.10	0
013 Fat (%)		0948	6.0900	0.06000	6.4767	0.49843	0.19555	53	-0.78	013.00	0
013 Fat (%)		2004	6.1500	0.10000	6.4767	0.49843	0.19555	53	-0.66	013.00	0
013 Fat (%)		0638	6.1500	0.10000	6.4767	0.49843	0.19555	53	-0.66	013.10	0
013 Fat (%)		0660	6.1500	0.12000	6.4767	0.49843	0.19555	53	-0.66	013.10	0
013 Fat (%)		0297	6.1550	0.03000	6.4767	0.49843	0.19555	53	-0.65	013.00	0
013 Fat (%)		0955	6.2100	0.00000	6.4767	0.49843	0.19555	53	-0.54	013.13	0
013 Fat (%)		0171	6.2150	0.17000	6.4767	0.49843	0.19555	53	-0.53	013.02	0
013 Fat (%)		0951	6.2950	0.07000	6.4767	0.49843	0.19555	53	-0.36	013.12	0
013 Fat (%)		2058	6.3000	0.08000	6.4767	0.49843	0.19555	53	-0.35	013.02	0
013 Fat (%)		0652	6.3000	0.00000	6.4767	0.49843	0.19555	53	-0.35	013.10	0
013 Fat (%)		0353	6.3050	0.05000	6.4767	0.49843	0.19555	53	-0.34	013.10	0
013 Fat (%)		0242	6.3200	0.52000	6.4767	0.49843	0.19555	53	-0.31	013.00	0
013 Fat (%)		0681	6.4150	0.05000	6.4767	0.49843	0.19555	53	-0.12	013.00	0
013 Fat (%)		0202	6.4650	0.13000	6.4767	0.49843	0.19555	53	-0.02	013.00	0
013 Fat (%)		0961	6.4800	0.30000	6.4767	0.49843	0.19555	53	0.01	013.12	0
013 Fat (%)		0164	6.4850	0.11000	6.4767	0.49843	0.19555	53	0.02	013.02	0
013 Fat (%)		0008	6.5250	0.37000	6.4767	0.49843	0.19555	53	0.10	013.02	0
013 Fat (%)		0695	6.5500	0.48000	6.4767	0.49843	0.19555	53	0.15	013.00	0
013 Fat (%)		0098	6.5800	0.18000	6.4767	0.49843	0.19555	53	0.21	013.02	0
013 Fat (%)		0208	6.6100	0.24000	6.4767	0.49843	0.19555	53	0.27	013.02	0
013 Fat (%)		0861	6.6350	0.39000	6.4767	0.49843	0.19555	53	0.32	013.00	0
013 Fat (%)		0051	6.6500	0.26000	6.4767	0.49843	0.19555	53	0.35	013.02	0
013 Fat (%)		0959	6.6550	0.09000	6.4767	0.49843	0.19555	53	0.36	013.02	0
013 Fat (%)		0650	6.6650	0.01000	6.4767	0.49843	0.19555	53	0.38	013.02	0
013 Fat (%)		0643	6.6850	0.11000	6.4767	0.49843	0.19555	53	0.42	013.02	0
013 Fat (%)		0921	6.6950	0.49000	6.4767	0.49843	0.19555	53	0.44	013.02	0
013 Fat (%)		0675	6.7050	0.03000	6.4767	0.49843	0.19555	53	0.46	013.02	0
013 Fat (%)		0618	6.7200	0.70000	6.4767	0.49843	0.19555	53	0.49	013.00	0
013 Fat (%)		0957	6.7400	0.36000	6.4767	0.49843	0.19555	53	0.53	013.02	0
013 Fat (%)		0229	6.7500	0.30000	6.4767	0.49843	0.19555	53	0.55	013.00	0
013 Fat (%)		0354	6.7600	0.02000	6.4767	0.49843	0.19555	53	0.57	013.02	0
013 Fat (%)		2023	6.8450	0.09000	6.4767	0.49843	0.19555	53	0.74	013.00	0
013 Fat (%)		0960	6.8900	0.04000	6.4767	0.49843	0.19555	53	0.83	013.02	0



Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
013	Fat (%)	0100	7.0550	0.21000	6.4767	0.49843	0.19555	53	1.16	013.02	0
013	Fat (%)	0504	7.0850	0.55000	6.4767	0.49843	0.19555	53	1.22	013.00	0
013	Fat (%)	0003	7.1250	0.19000	6.4767	0.49843	0.19555	53	1.30	013.02	0
013	Fat (%)	0682	7.1400	0.10000	6.4767	0.49843	0.19555	53	1.33	013.02	0
013	Fat (%)	0337	7.1500	0.68000	6.4767	0.49843	0.19555	53	1.35	013.02	0
013	Fat (%)	0968	7.2290	0.07400	6.4767	0.49843	0.19555	53	1.51	013.10	0
013	Fat (%)	0553	7.2300	0.20000	6.4767	0.49843	0.19555	53	1.51	013.02	0
013	Fat (%)	0265	7.7000	0.20000	6.4767	0.49843	0.19555	53	2.45	013.13	0
013	Fat (%)	0581	7.8600	0.60000	6.4767	0.49843	0.19555	53	2.78	013.13	0
013	Fat (%)	2016	6.5650	1.4100	6.4767	0.49843	0.19555	53	0.18	013.02	1
013	Fat (%)	0910	0.00000	0.00000	6.4767	0.49843	0.19555	53	-12.99	013.00	2
015	Aluminum (mg / kg (ppm))	0297	88.500	5.0000	107.05	17.304	4.6445	14	-1.07	015.43	0
015	Aluminum (mg / kg (ppm))	0510	89.000	0.00000	107.05	17.304	4.6445	14	-1.04	015.43	0
015	Aluminum (mg / kg (ppm))	2058	91.015	6.6900	107.05	17.304	4.6445	14	-0.93	015.43	0
015	Aluminum (mg / kg (ppm))	0353	91.900	4.0800	107.05	17.304	4.6445	14	-0.88	015.43	0
015	Aluminum (mg / kg (ppm))	0049	94.120	3.1800	107.05	17.304	4.6445	14	-0.75	015.41	0
015	Aluminum (mg / kg (ppm))	0011	100.95	0.02250	107.05	17.304	4.6445	14	-0.35	015.41	0
015	Aluminum (mg / kg (ppm))	0164	103.00	0.00000	107.05	17.304	4.6445	14	-0.23	015.41	0
015	Aluminum (mg / kg (ppm))	0560	110.75	0.50000	107.05	17.304	4.6445	14	0.21	015.42	0
015	Aluminum (mg / kg (ppm))	0918	112.53	20.540	107.05	17.304	4.6445	14	0.32	015.43	0
015	Aluminum (mg / kg (ppm))	0171	113.50	7.0000	107.05	17.304	4.6445	14	0.37	015.41	0
015	Aluminum (mg / kg (ppm))	0345	116.50	3.0000	107.05	17.304	4.6445	14	0.55	015.43	0
015	Aluminum (mg / kg (ppm))	0553	125.00	6.0000	107.05	17.304	4.6445	14	1.04	015.53	0
015	Aluminum (mg / kg (ppm))	0964	128.94	1.0100	107.05	17.304	4.6445	14	1.26	015.41	0
015	Aluminum (mg / kg (ppm))	0520	136.00	8.0000	107.05	17.304	4.6445	14	1.67	015.41	0
015	Aluminum (mg / kg (ppm))	0169	256.00	34.000	107.05	17.304	4.6445	14	8.61	015.43	1
017	Boron (mg / kg (ppm))	0297	2.5000	1.0000	5.1254	1.0331	0.38083	12	-2.54	017.43	0
017	Boron (mg / kg (ppm))	0510	4.0000	0.00000	5.1254	1.0331	0.38083	12	-1.09	017.43	0
017	Boron (mg / kg (ppm))	0693	4.2400	0.98000	5.1254	1.0331	0.38083	12	-0.86	017.42	0
017	Boron (mg / kg (ppm))	0358	4.6850	0.15000	5.1254	1.0331	0.38083	12	-0.43	017.41	0
017	Boron (mg / kg (ppm))	0560	4.9350	0.07000	5.1254	1.0331	0.38083	12	-0.18	017.42	0
017	Boron (mg / kg (ppm))	2055	5.1000	0.20000	5.1254	1.0331	0.38083	12	-0.02	017.99	0
017	Boron (mg / kg (ppm))	0918	5.1700	0.14000	5.1254	1.0331	0.38083	12	0.04	017.43	0
017	Boron (mg / kg (ppm))	0345	5.2500	0.10000	5.1254	1.0331	0.38083	12	0.12	017.43	0
017	Boron (mg / kg (ppm))	0045	5.5000	0.60000	5.1254	1.0331	0.38083	12	0.36	017.42	0
017	Boron (mg / kg (ppm))	0353	6.1700	0.14000	5.1254	1.0331	0.38083	12	1.01	017.43	0
017	Boron (mg / kg (ppm))	0049	6.1850	1.1500	5.1254	1.0331	0.38083	12	1.03	017.41	0
017	Boron (mg / kg (ppm))	0294	7.2200	0.04000	5.1254	1.0331	0.38083	12	2.03	017.42	0
019	Calcium (%)	0536	0.66000	0.06000	2.0829	0.09613	0.04273	162	-14.80	019.31	0
019	Calcium (%)	0638	1.0400	0.00000	2.0829	0.09613	0.04273	162	-10.85	019.31	0
019	Calcium (%)	0510	1.7150	0.01000	2.0829	0.09613	0.04273	162	-3.83	019.43	0
019	Calcium (%)	0852	1.7500	0.10000	2.0829	0.09613	0.04273	162	-3.46	019.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
019	Calcium (%)	0047	1.7550	0.11000	2.0829	0.09613	0.04273	162	-3.41	019.52	0
019	Calcium (%)	0552	1.7900	0.00000	2.0829	0.09613	0.04273	162	-3.05	019.00	0
019	Calcium (%)	0154	1.8011	0.02460	2.0829	0.09613	0.04273	162	-2.93	019.52	0
019	Calcium (%)	0405	1.8250	0.03000	2.0829	0.09613	0.04273	162	-2.68	019.41	0
019	Calcium (%)	2058	1.8400	0.02000	2.0829	0.09613	0.04273	162	-2.53	019.43	0
019	Calcium (%)	0089	1.8500	0.00000	2.0829	0.09613	0.04273	162	-2.42	019.41	0
019	Calcium (%)	0852	1.8500	0.10000	2.0829	0.09613	0.04273	162	-2.42	019.99	0
019	Calcium (%)	2006	1.8560	0.07200	2.0829	0.09613	0.04273	162	-2.36	019.09	0
019	Calcium (%)	0043	1.8700	0.04000	2.0829	0.09613	0.04273	162	-2.21	019.43	0
019	Calcium (%)	0108	1.8850	0.01000	2.0829	0.09613	0.04273	162	-2.06	019.31	0
019	Calcium (%)	0144	1.8950	0.01000	2.0829	0.09613	0.04273	162	-1.95	019.41	0
019	Calcium (%)	2006	1.8980	0.03200	2.0829	0.09613	0.04273	162	-1.92	019.00	0
019	Calcium (%)	0848	1.9100	0.02000	2.0829	0.09613	0.04273	162	-1.80	019.41	0
019	Calcium (%)	0889	1.9500	0.01200	2.0829	0.09613	0.04273	162	-1.38	019.99	0
019	Calcium (%)	0142	1.9550	0.00000	2.0829	0.09613	0.04273	162	-1.33	019.31	0
019	Calcium (%)	0948	1.9550	0.01000	2.0829	0.09613	0.04273	162	-1.33	019.31	0
019	Calcium (%)	0139	1.9590	0.04800	2.0829	0.09613	0.04273	162	-1.29	019.31	0
019	Calcium (%)	0623	1.9751	0.25260	2.0829	0.09613	0.04273	162	-1.12	019.00	0
019	Calcium (%)	0878	1.9800	0.02000	2.0829	0.09613	0.04273	162	-1.07	019.00	0
019	Calcium (%)	0553	1.9800	0.06000	2.0829	0.09613	0.04273	162	-1.07	019.41	0
019	Calcium (%)	0693	1.9800	0.02000	2.0829	0.09613	0.04273	162	-1.07	019.42	0
019	Calcium (%)	0242	1.9850	0.15000	2.0829	0.09613	0.04273	162	-1.02	019.99	0
019	Calcium (%)	0018	1.9900	0.02000	2.0829	0.09613	0.04273	162	-0.97	019.41	0
019	Calcium (%)	0891	1.9921	0.00720	2.0829	0.09613	0.04273	162	-0.94	019.00	0
019	Calcium (%)	0171	1.9950	0.05000	2.0829	0.09613	0.04273	162	-0.91	019.41	0
019	Calcium (%)	0358	1.9950	0.09000	2.0829	0.09613	0.04273	162	-0.91	019.41	0
019	Calcium (%)	0903	2.0000	0.02000	2.0829	0.09613	0.04273	162	-0.86	019.00	0
019	Calcium (%)	0194	2.0050	0.01000	2.0829	0.09613	0.04273	162	-0.81	019.00	0
019	Calcium (%)	0353	2.0050	0.09000	2.0829	0.09613	0.04273	162	-0.81	019.43	0
019	Calcium (%)	0629	2.0100	0.02000	2.0829	0.09613	0.04273	162	-0.76	019.08	0
019	Calcium (%)	0036	2.0117	0.09940	2.0829	0.09613	0.04273	162	-0.74	019.43	0
019	Calcium (%)	0178	2.0150	0.01000	2.0829	0.09613	0.04273	162	-0.71	019.31	0
019	Calcium (%)	2018	2.0150	0.01000	2.0829	0.09613	0.04273	162	-0.71	019.31	0
019	Calcium (%)	0881	2.0165	0.03700	2.0829	0.09613	0.04273	162	-0.69	019.00	0
019	Calcium (%)	0937	2.0200	0.00000	2.0829	0.09613	0.04273	162	-0.65	019.00	0
019	Calcium (%)	0893	2.0250	0.01000	2.0829	0.09613	0.04273	162	-0.60	019.00	0
019	Calcium (%)	0233	2.0250	0.01000	2.0829	0.09613	0.04273	162	-0.60	019.31	0
019	Calcium (%)	0011	2.0276	0.02690	2.0829	0.09613	0.04273	162	-0.58	019.41	0
019	Calcium (%)	0563	2.0333	0.06300	2.0829	0.09613	0.04273	162	-0.52	019.31	0
019	Calcium (%)	0042	2.0350	0.09000	2.0829	0.09613	0.04273	162	-0.50	019.42	0
019	Calcium (%)	0692	2.0350	0.01000	2.0829	0.09613	0.04273	162	-0.50	019.42	0
019	Calcium (%)	2067	2.0400	0.00000	2.0829	0.09613	0.04273	162	-0.45	019.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
019	Calcium (%)	2068	2.0400	0.00000	2.0829	0.09613	0.04273	162	-0.45	019.00	0
019	Calcium (%)	0939	2.0400	0.24000	2.0829	0.09613	0.04273	162	-0.45	019.31	0
019	Calcium (%)	0074	2.0400	0.02000	2.0829	0.09613	0.04273	162	-0.45	019.41	0
019	Calcium (%)	0653	2.0410	0.00000	2.0829	0.09613	0.04273	162	-0.44	019.31	0
019	Calcium (%)	0590	2.0450	0.07000	2.0829	0.09613	0.04273	162	-0.39	019.08	0
019	Calcium (%)	0910	2.0450	0.09000	2.0829	0.09613	0.04273	162	-0.39	019.41	0
019	Calcium (%)	0065	2.0488	0.00350	2.0829	0.09613	0.04273	162	-0.36	019.31	0
019	Calcium (%)	0504	2.0495	0.01700	2.0829	0.09613	0.04273	162	-0.35	019.42	0
019	Calcium (%)	0098	2.0500	0.02000	2.0829	0.09613	0.04273	162	-0.34	019.41	0
019	Calcium (%)	0878	2.0500	0.10000	2.0829	0.09613	0.04273	162	-0.34	019.41	0
019	Calcium (%)	0278	2.0500	0.16000	2.0829	0.09613	0.04273	162	-0.34	019.42	0
019	Calcium (%)	0629	2.0500	0.00000	2.0829	0.09613	0.04273	162	-0.34	019.43	0
019	Calcium (%)	2067	2.0510	0.05800	2.0829	0.09613	0.04273	162	-0.33	019.31	0
019	Calcium (%)	0723	2.0525	0.00000	2.0829	0.09613	0.04273	162	-0.32	019.31	0
019	Calcium (%)	0620	2.0540	0.03770	2.0829	0.09613	0.04273	162	-0.30	019.00	0
019	Calcium (%)	0051	2.0549	0.00140	2.0829	0.09613	0.04273	162	-0.29	019.41	0
019	Calcium (%)	0683	2.0550	0.03000	2.0829	0.09613	0.04273	162	-0.29	019.00	0
019	Calcium (%)	0045	2.0550	0.07000	2.0829	0.09613	0.04273	162	-0.29	019.42	0
019	Calcium (%)	0265	2.0550	0.11000	2.0829	0.09613	0.04273	162	-0.29	019.42	0
019	Calcium (%)	0874	2.0560	0.01800	2.0829	0.09613	0.04273	162	-0.28	019.31	0
019	Calcium (%)	0849	2.0565	0.00700	2.0829	0.09613	0.04273	162	-0.27	019.08	0
019	Calcium (%)	0897	2.0600	0.02000	2.0829	0.09613	0.04273	162	-0.24	019.00	0
019	Calcium (%)	2022	2.0600	0.00000	2.0829	0.09613	0.04273	162	-0.24	019.31	0
019	Calcium (%)	0205	2.0600	0.04000	2.0829	0.09613	0.04273	162	-0.24	019.42	0
019	Calcium (%)	0043	2.0650	0.07000	2.0829	0.09613	0.04273	162	-0.19	019.00	0
019	Calcium (%)	0407	2.0650	0.05000	2.0829	0.09613	0.04273	162	-0.19	019.41	0
019	Calcium (%)	0297	2.0650	0.01000	2.0829	0.09613	0.04273	162	-0.19	019.43	0
019	Calcium (%)	0138	2.0685	0.07700	2.0829	0.09613	0.04273	162	-0.15	019.08	0
019	Calcium (%)	0921	2.0710	0.05000	2.0829	0.09613	0.04273	162	-0.12	019.31	0
019	Calcium (%)	0035	2.0715	0.02100	2.0829	0.09613	0.04273	162	-0.12	019.42	0
019	Calcium (%)	2063	2.0750	0.01000	2.0829	0.09613	0.04273	162	-0.08	019.08	0
019	Calcium (%)	0650	2.0750	0.05000	2.0829	0.09613	0.04273	162	-0.08	019.31	0
019	Calcium (%)	0017	2.0750	0.03000	2.0829	0.09613	0.04273	162	-0.08	019.43	0
019	Calcium (%)	0298	2.0800	0.02000	2.0829	0.09613	0.04273	162	-0.03	019.41	0
019	Calcium (%)	2012	2.0800	0.02000	2.0829	0.09613	0.04273	162	-0.03	019.41	0
019	Calcium (%)	0199	2.0800	0.04000	2.0829	0.09613	0.04273	162	-0.03	019.43	0
019	Calcium (%)	0651	2.0815	0.00100	2.0829	0.09613	0.04273	162	-0.01	019.00	0
019	Calcium (%)	0038	2.0850	0.03000	2.0829	0.09613	0.04273	162	0.02	019.31	0
019	Calcium (%)	0019	2.0850	0.01000	2.0829	0.09613	0.04273	162	0.02	019.41	0
019	Calcium (%)	0610	2.0850	0.03000	2.0829	0.09613	0.04273	162	0.02	019.41	0
019	Calcium (%)	0168	2.0850	0.05000	2.0829	0.09613	0.04273	162	0.02	019.43	0
019	Calcium (%)	0345	2.0850	0.05000	2.0829	0.09613	0.04273	162	0.02	019.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
019	Calcium (%)	2062	2.0950	0.06550	2.0829	0.09613	0.04273	162	0.13	019.00	0
019	Calcium (%)	0049	2.0950	0.03000	2.0829	0.09613	0.04273	162	0.13	019.41	0
019	Calcium (%)	0685	2.0950	0.01000	2.0829	0.09613	0.04273	162	0.13	019.44	0
019	Calcium (%)	2065	2.1000	0.00660	2.0829	0.09613	0.04273	162	0.18	019.00	0
019	Calcium (%)	0294	2.1000	0.00000	2.0829	0.09613	0.04273	162	0.18	019.42	0
019	Calcium (%)	0008	2.1015	0.08500	2.0829	0.09613	0.04273	162	0.19	019.43	0
019	Calcium (%)	0263	2.1024	0.00160	2.0829	0.09613	0.04273	162	0.20	019.32	0
019	Calcium (%)	0208	2.1045	0.06900	2.0829	0.09613	0.04273	162	0.22	019.41	0
019	Calcium (%)	0895	2.1050	0.01000	2.0829	0.09613	0.04273	162	0.23	019.00	0
019	Calcium (%)	0689	2.1050	0.03000	2.0829	0.09613	0.04273	162	0.23	019.31	0
019	Calcium (%)	0868	2.1050	0.01000	2.0829	0.09613	0.04273	162	0.23	019.31	0
019	Calcium (%)	0300	2.1055	0.05100	2.0829	0.09613	0.04273	162	0.24	019.41	0
019	Calcium (%)	0885	2.1100	0.00000	2.0829	0.09613	0.04273	162	0.28	019.08	0
019	Calcium (%)	0208	2.1100	0.04000	2.0829	0.09613	0.04273	162	0.28	019.31	0
019	Calcium (%)	0357	2.1100	0.04000	2.0829	0.09613	0.04273	162	0.28	019.42	0
019	Calcium (%)	0354	2.1100	0.02000	2.0829	0.09613	0.04273	162	0.28	019.31	0
019	Calcium (%)	0187	2.1100	0.02000	2.0829	0.09613	0.04273	162	0.28	019.42	0
019	Calcium (%)	0968	2.1140	0.01200	2.0829	0.09613	0.04273	162	0.32	019.43	0
019	Calcium (%)	0038	2.1150	0.03000	2.0829	0.09613	0.04273	162	0.33	019.43	0
019	Calcium (%)	2011	2.1200	0.04000	2.0829	0.09613	0.04273	162	0.39	019.00	0
019	Calcium (%)	0003	2.1200	0.02000	2.0829	0.09613	0.04273	162	0.39	019.41	0
019	Calcium (%)	0675	2.1200	0.00000	2.0829	0.09613	0.04273	162	0.39	019.43	0
019	Calcium (%)	2004	2.1200	0.04000	2.0829	0.09613	0.04273	162	0.39	019.44	0
019	Calcium (%)	0148	2.1250	0.01000	2.0829	0.09613	0.04273	162	0.44	019.41	0
019	Calcium (%)	0037	2.1250	0.01000	2.0829	0.09613	0.04273	162	0.44	019.43	0
019	Calcium (%)	0504	2.1255	0.18100	2.0829	0.09613	0.04273	162	0.44	019.32	0
019	Calcium (%)	0014	2.1300	0.04000	2.0829	0.09613	0.04273	162	0.49	019.31	0
019	Calcium (%)	0229	2.1350	0.01000	2.0829	0.09613	0.04273	162	0.54	019.41	0
019	Calcium (%)	0033	2.1350	0.07000	2.0829	0.09613	0.04273	162	0.54	019.43	0
019	Calcium (%)	0560	2.1362	0.03350	2.0829	0.09613	0.04273	162	0.55	019.42	0
019	Calcium (%)	0164	2.1400	0.00000	2.0829	0.09613	0.04273	162	0.59	019.41	0
019	Calcium (%)	0505	2.1450	0.01000	2.0829	0.09613	0.04273	162	0.65	019.31	0
019	Calcium (%)	0001	2.1490	0.01400	2.0829	0.09613	0.04273	162	0.69	019.31	0
019	Calcium (%)	0695	2.1500	0.02000	2.0829	0.09613	0.04273	162	0.70	019.41	0
019	Calcium (%)	2006	2.1500	0.06000	2.0829	0.09613	0.04273	162	0.70	019.99	0
019	Calcium (%)	0918	2.1525	0.08700	2.0829	0.09613	0.04273	162	0.72	019.43	0
019	Calcium (%)	0687	2.1550	0.05000	2.0829	0.09613	0.04273	162	0.75	019.31	0
019	Calcium (%)	0035	2.1570	0.07200	2.0829	0.09613	0.04273	162	0.77	019.32	0
019	Calcium (%)	0619	2.1600	0.06000	2.0829	0.09613	0.04273	162	0.80	019.41	0
019	Calcium (%)	0681	2.1650	0.01000	2.0829	0.09613	0.04273	162	0.85	019.00	0
019	Calcium (%)	0609	2.1650	0.09000	2.0829	0.09613	0.04273	162	0.85	019.32	0
019	Calcium (%)	0169	2.1700	0.02000	2.0829	0.09613	0.04273	162	0.91	019.32	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
019	Calcium (%)	2053	2.1700	0.00000	2.0829	0.09613	0.04273	162	0.91	019.42	0
019	Calcium (%)	0100	2.1750	0.03000	2.0829	0.09613	0.04273	162	0.96	019.41	0
019	Calcium (%)	0083	2.1800	0.06000	2.0829	0.09613	0.04273	162	1.01	019.41	0
019	Calcium (%)	2023	2.1800	0.02000	2.0829	0.09613	0.04273	162	1.01	019.44	0
019	Calcium (%)	2042	2.1800	0.00000	2.0829	0.09613	0.04273	162	1.01	019.08	0
019	Calcium (%)	0674	2.1800	0.00000	2.0829	0.09613	0.04273	162	1.01	019.31	0
019	Calcium (%)	0682	2.1800	0.00000	2.0829	0.09613	0.04273	162	1.01	019.41	0
019	Calcium (%)	0726	2.1850	0.00800	2.0829	0.09613	0.04273	162	1.06	019.42	0
019	Calcium (%)	0689	2.1950	0.05000	2.0829	0.09613	0.04273	162	1.17	019.08	0
019	Calcium (%)	0612	2.1950	0.01000	2.0829	0.09613	0.04273	162	1.17	019.32	0
019	Calcium (%)	0036	2.2031	0.02930	2.0829	0.09613	0.04273	162	1.25	019.03	0
019	Calcium (%)	0511	2.2050	0.07000	2.0829	0.09613	0.04273	162	1.27	019.41	0
019	Calcium (%)	0175	2.2100	0.02000	2.0829	0.09613	0.04273	162	1.32	019.31	0
019	Calcium (%)	0720	2.2100	0.14000	2.0829	0.09613	0.04273	162	1.32	019.41	0
019	Calcium (%)	0598	2.2115	0.05500	2.0829	0.09613	0.04273	162	1.34	019.41	0
019	Calcium (%)	0572	2.2235	0.06500	2.0829	0.09613	0.04273	162	1.46	019.53	0
019	Calcium (%)	0425	2.2250	0.05000	2.0829	0.09613	0.04273	162	1.48	019.43	0
019	Calcium (%)	0504	2.2295	0.09700	2.0829	0.09613	0.04273	162	1.53	019.33	0
019	Calcium (%)	0964	2.2326	0.22400	2.0829	0.09613	0.04273	162	1.56	019.43	0
019	Calcium (%)	0010	2.2350	0.07000	2.0829	0.09613	0.04273	162	1.58	019.33	0
019	Calcium (%)	0098	2.2350	0.11000	2.0829	0.09613	0.04273	162	1.58	019.43	0
019	Calcium (%)	0202	2.2400	0.02000	2.0829	0.09613	0.04273	162	1.63	019.42	0
019	Calcium (%)	0964	2.2446	0.00420	2.0829	0.09613	0.04273	162	1.68	019.41	0
019	Calcium (%)	0555	2.2495	0.04490	2.0829	0.09613	0.04273	162	1.73	019.42	0
019	Calcium (%)	0413	2.2500	0.10000	2.0829	0.09613	0.04273	162	1.74	019.41	0
019	Calcium (%)	0226	2.2700	0.00000	2.0829	0.09613	0.04273	162	1.95	019.41	0
019	Calcium (%)	0106	2.2800	0.00000	2.0829	0.09613	0.04273	162	2.05	019.42	0
019	Calcium (%)	0028	2.2850	0.07000	2.0829	0.09613	0.04273	162	2.10	019.43	0
019	Calcium (%)	0152	2.3500	0.10000	2.0829	0.09613	0.04273	162	2.78	019.31	0
019	Calcium (%)	0520	2.4200	0.08000	2.0829	0.09613	0.04273	162	3.51	019.41	0
019	Calcium (%)	0066	2.4650	0.13000	2.0829	0.09613	0.04273	162	3.97	019.31	0
019	Calcium (%)	0951	2.8780	0.17000	2.0829	0.09613	0.04273	162	8.27	019.31	0
019	Calcium (%)	0676	1.5300	0.92000	2.0829	0.09613	0.04273	162	-5.75	019.99	1
019	Calcium (%)	2055	1.6840	0.30800	2.0829	0.09613	0.04273	162	-4.15	019.99	1
019	Calcium (%)	0337	2.8200	0.30000	2.0829	0.09613	0.04273	162	7.67	019.00	1
019	Calcium (%)	0190	0.58000	0.04000	2.0829	0.09613	0.04273	162	-15.63	019.42	2
021	Cobalt (mg / kg (ppm))	0619	0.85750	0.00100	1.7130	0.55109	0.17000	25	-1.55	021.41	0
021	Cobalt (mg / kg (ppm))	0098	0.93100	0.04200	1.7130	0.55109	0.17000	25	-1.42	021.52	0
021	Cobalt (mg / kg (ppm))	0560	0.99050	0.34900	1.7130	0.55109	0.17000	25	-1.31	021.42	0
021	Cobalt (mg / kg (ppm))	2055	1.0000	0.40000	1.7130	0.55109	0.17000	25	-1.29	021.99	0
021	Cobalt (mg / kg (ppm))	0693	1.1100	0.20000	1.7130	0.55109	0.17000	25	-1.09	021.00	0
021	Cobalt (mg / kg (ppm))	0918	1.2250	0.63000	1.7130	0.55109	0.17000	25	-0.89	021.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
021	Cobalt (mg / kg (ppm))	0345	1.2850	0.13000	1.7130	0.55109	0.17000	25	-0.78	021.43	0
021	Cobalt (mg / kg (ppm))	0171	1.4350	0.09000	1.7130	0.55109	0.17000	25	-0.50	021.41	0
021	Cobalt (mg / kg (ppm))	0169	1.5400	0.02000	1.7130	0.55109	0.17000	25	-0.31	021.43	0
021	Cobalt (mg / kg (ppm))	0164	1.6000	0.20000	1.7130	0.55109	0.17000	25	-0.21	021.31	0
021	Cobalt (mg / kg (ppm))	0689	1.6000	0.00000	1.7130	0.55109	0.17000	25	-0.21	021.31	0
021	Cobalt (mg / kg (ppm))	0011	1.7310	0.09800	1.7130	0.55109	0.17000	25	0.03	021.41	0
021	Cobalt (mg / kg (ppm))	0168	1.7650	0.45000	1.7130	0.55109	0.17000	25	0.09	021.53	0
021	Cobalt (mg / kg (ppm))	0510	1.8450	0.01000	1.7130	0.55109	0.17000	25	0.24	021.43	0
021	Cobalt (mg / kg (ppm))	0553	1.8500	0.42000	1.7130	0.55109	0.17000	25	0.25	021.53	0
021	Cobalt (mg / kg (ppm))	0043	1.9600	0.30000	1.7130	0.55109	0.17000	25	0.45	021.43	0
021	Cobalt (mg / kg (ppm))	0939	2.0150	0.03000	1.7130	0.55109	0.17000	25	0.55	021.31	0
021	Cobalt (mg / kg (ppm))	0199	2.0300	0.16000	1.7130	0.55109	0.17000	25	0.58	021.53	0
021	Cobalt (mg / kg (ppm))	0038	2.0500	0.10000	1.7130	0.55109	0.17000	25	0.61	021.43	0
021	Cobalt (mg / kg (ppm))	0610	2.1150	0.11000	1.7130	0.55109	0.17000	25	0.73	021.34	0
021	Cobalt (mg / kg (ppm))	0964	2.1210	0.05970	1.7130	0.55109	0.17000	25	0.74	021.43	0
021	Cobalt (mg / kg (ppm))	0964	2.2453	0.11020	1.7130	0.55109	0.17000	25	0.97	021.41	0
021	Cobalt (mg / kg (ppm))	0003	2.3000	0.28000	1.7130	0.55109	0.17000	25	1.07	021.43	0
021	Cobalt (mg / kg (ppm))	0106	2.5000	0.06000	1.7130	0.55109	0.17000	25	1.43	021.42	0
021	Cobalt (mg / kg (ppm))	2022	3.0000	0.00000	1.7130	0.55109	0.17000	25	2.34	021.31	0
022	Copper (mg / kg (ppm))	0695	2.0800	0.10000	14.093	2.7236	0.71114	93	-4.41	022.41	0
022	Copper (mg / kg (ppm))	0610	2.1150	0.11000	14.093	2.7236	0.71114	93	-4.40	022.41	0
022	Copper (mg / kg (ppm))	0653	8.8780	0.12200	14.093	2.7236	0.71114	93	-1.91	022.31	0
022	Copper (mg / kg (ppm))	0504	9.4815	2.7570	14.093	2.7236	0.71114	93	-1.69	022.33	0
022	Copper (mg / kg (ppm))	0590	9.5000	1.0000	14.093	2.7236	0.71114	93	-1.69	022.31	0
022	Copper (mg / kg (ppm))	0511	9.5000	1.0000	14.093	2.7236	0.71114	93	-1.69	022.41	0
022	Copper (mg / kg (ppm))	0619	9.6300	0.94000	14.093	2.7236	0.71114	93	-1.64	022.41	0
022	Copper (mg / kg (ppm))	0049	9.7900	0.76000	14.093	2.7236	0.71114	93	-1.58	022.41	0
022	Copper (mg / kg (ppm))	0629	10.000	0.00000	14.093	2.7236	0.71114	93	-1.50	022.31	0
022	Copper (mg / kg (ppm))	0003	10.500	1.0000	14.093	2.7236	0.71114	93	-1.32	022.41	0
022	Copper (mg / kg (ppm))	0358	10.635	0.33000	14.093	2.7236	0.71114	93	-1.27	022.41	0
022	Copper (mg / kg (ppm))	0229	11.000	0.00000	14.093	2.7236	0.71114	93	-1.14	022.41	0
022	Copper (mg / kg (ppm))	0553	11.000	0.20000	14.093	2.7236	0.71114	93	-1.14	022.41	0
022	Copper (mg / kg (ppm))	0011	11.043	0.30500	14.093	2.7236	0.71114	93	-1.12	022.41	0
022	Copper (mg / kg (ppm))	0638	11.100	1.0000	14.093	2.7236	0.71114	93	-1.10	022.31	0
022	Copper (mg / kg (ppm))	0598	11.145	1.2500	14.093	2.7236	0.71114	93	-1.08	022.41	0
022	Copper (mg / kg (ppm))	0171	11.185	0.51000	14.093	2.7236	0.71114	93	-1.07	022.41	0
022	Copper (mg / kg (ppm))	0921	11.490	1.3000	14.093	2.7236	0.71114	93	-0.96	022.31	0
022	Copper (mg / kg (ppm))	0878	11.500	1.0000	14.093	2.7236	0.71114	93	-0.95	022.41	0
022	Copper (mg / kg (ppm))	0199	11.500	0.20000	14.093	2.7236	0.71114	93	-0.95	022.53	0
022	Copper (mg / kg (ppm))	0300	11.831	0.90400	14.093	2.7236	0.71114	93	-0.83	022.41	0
022	Copper (mg / kg (ppm))	0066	11.985	0.09000	14.093	2.7236	0.71114	93	-0.77	022.31	0
022	Copper (mg / kg (ppm))	2018	12.100	0.72000	14.093	2.7236	0.71114	93	-0.73	022.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
022	Copper (mg / kg (ppm))	0098	12.150	0.54000	14.093	2.7236	0.71114	93	-0.71	022.41	0
022	Copper (mg / kg (ppm))	0723	12.500	0.00000	14.093	2.7236	0.71114	93	-0.58	022.31	0
022	Copper (mg / kg (ppm))	0148	12.500	0.40000	14.093	2.7236	0.71114	93	-0.58	022.41	0
022	Copper (mg / kg (ppm))	0164	12.500	1.0000	14.093	2.7236	0.71114	93	-0.58	022.41	0
022	Copper (mg / kg (ppm))	0629	12.500	0.00000	14.093	2.7236	0.71114	93	-0.58	022.43	0
022	Copper (mg / kg (ppm))	0505	12.600	0.40000	14.093	2.7236	0.71114	93	-0.55	022.31	0
022	Copper (mg / kg (ppm))	0563	12.605	0.45000	14.093	2.7236	0.71114	93	-0.55	022.31	0
022	Copper (mg / kg (ppm))	0674	12.725	1.2700	14.093	2.7236	0.71114	93	-0.50	022.31	0
022	Copper (mg / kg (ppm))	0948	12.825	0.17000	14.093	2.7236	0.71114	93	-0.47	022.33	0
022	Copper (mg / kg (ppm))	0242	13.000	2.0000	14.093	2.7236	0.71114	93	-0.40	022.99	0
022	Copper (mg / kg (ppm))	0208	13.005	0.69000	14.093	2.7236	0.71114	93	-0.40	022.41	0
022	Copper (mg / kg (ppm))	0939	13.205	0.09000	14.093	2.7236	0.71114	93	-0.33	022.31	0
022	Copper (mg / kg (ppm))	0353	13.335	0.43000	14.093	2.7236	0.71114	93	-0.28	022.43	0
022	Copper (mg / kg (ppm))	0354	13.350	0.50000	14.093	2.7236	0.71114	93	-0.27	022.31	0
022	Copper (mg / kg (ppm))	0868	13.350	0.10000	14.093	2.7236	0.71114	93	-0.27	022.31	0
022	Copper (mg / kg (ppm))	0951	13.350	2.9000	14.093	2.7236	0.71114	93	-0.27	022.31	0
022	Copper (mg / kg (ppm))	2004	13.550	1.1000	14.093	2.7236	0.71114	93	-0.20	022.44	0
022	Copper (mg / kg (ppm))	0720	13.680	1.1090	14.093	2.7236	0.71114	93	-0.15	022.41	0
022	Copper (mg / kg (ppm))	0407	13.865	0.99000	14.093	2.7236	0.71114	93	-0.08	022.41	0
022	Copper (mg / kg (ppm))	0964	13.936	0.69600	14.093	2.7236	0.71114	93	-0.06	022.41	0
022	Copper (mg / kg (ppm))	0074	14.000	0.00000	14.093	2.7236	0.71114	93	-0.03	022.41	0
022	Copper (mg / kg (ppm))	0520	14.000	0.00000	14.093	2.7236	0.71114	93	-0.03	022.41	0
022	Copper (mg / kg (ppm))	0910	14.000	0.00000	14.093	2.7236	0.71114	93	-0.03	022.41	0
022	Copper (mg / kg (ppm))	0051	14.240	0.40000	14.093	2.7236	0.71114	93	0.05	022.41	0
022	Copper (mg / kg (ppm))	0692	14.450	0.50000	14.093	2.7236	0.71114	93	0.13	022.42	0
022	Copper (mg / kg (ppm))	0265	14.500	3.0000	14.093	2.7236	0.71114	93	0.15	022.42	0
022	Copper (mg / kg (ppm))	0510	14.500	1.0000	14.093	2.7236	0.71114	93	0.15	022.43	0
022	Copper (mg / kg (ppm))	0178	14.550	0.30000	14.093	2.7236	0.71114	93	0.17	022.31	0
022	Copper (mg / kg (ppm))	0345	14.650	0.10000	14.093	2.7236	0.71114	93	0.20	022.43	0
022	Copper (mg / kg (ppm))	0918	14.780	0.02000	14.093	2.7236	0.71114	93	0.25	022.43	0
022	Copper (mg / kg (ppm))	0689	14.800	0.40000	14.093	2.7236	0.71114	93	0.26	022.31	0
022	Copper (mg / kg (ppm))	0425	14.950	0.10000	14.093	2.7236	0.71114	93	0.31	022.43	0
022	Copper (mg / kg (ppm))	0004	15.000	0.00000	14.093	2.7236	0.71114	93	0.33	022.41	0
022	Copper (mg / kg (ppm))	0226	15.000	0.00000	14.093	2.7236	0.71114	93	0.33	022.41	0
022	Copper (mg / kg (ppm))	0297	15.000	0.00000	14.093	2.7236	0.71114	93	0.33	022.43	0
022	Copper (mg / kg (ppm))	0968	15.000	0.00000	14.093	2.7236	0.71114	93	0.33	022.43	0
022	Copper (mg / kg (ppm))	0045	15.200	1.0000	14.093	2.7236	0.71114	93	0.41	022.42	0
022	Copper (mg / kg (ppm))	0008	15.210	0.82000	14.093	2.7236	0.71114	93	0.41	022.43	0
022	Copper (mg / kg (ppm))	0682	15.300	0.00000	14.093	2.7236	0.71114	93	0.44	022.41	0
022	Copper (mg / kg (ppm))	0038	15.350	0.90000	14.093	2.7236	0.71114	93	0.46	022.43	0
022	Copper (mg / kg (ppm))	2023	15.450	0.50000	14.093	2.7236	0.71114	93	0.50	022.53	0
022	Copper (mg / kg (ppm))	0555	15.459	0.25330	14.093	2.7236	0.71114	93	0.50	022.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
022	Copper (mg / kg (ppm))	0100	15.500	1.0000	14.093	2.7236	0.71114	93	0.52	022.41	0
022	Copper (mg / kg (ppm))	0405	15.500	1.0000	14.093	2.7236	0.71114	93	0.52	022.41	0
022	Copper (mg / kg (ppm))	2055	15.600	0.80000	14.093	2.7236	0.71114	93	0.55	022.99	0
022	Copper (mg / kg (ppm))	0294	15.620	0.04000	14.093	2.7236	0.71114	93	0.56	022.42	0
022	Copper (mg / kg (ppm))	0560	15.700	0.20000	14.093	2.7236	0.71114	93	0.59	022.42	0
022	Copper (mg / kg (ppm))	0413	15.800	0.20000	14.093	2.7236	0.71114	93	0.63	022.42	0
022	Copper (mg / kg (ppm))	0357	16.000	0.00000	14.093	2.7236	0.71114	93	0.70	022.42	0
022	Copper (mg / kg (ppm))	0675	16.015	0.25000	14.093	2.7236	0.71114	93	0.71	022.43	0
022	Copper (mg / kg (ppm))	0504	16.060	0.68000	14.093	2.7236	0.71114	93	0.72	022.32	0
022	Copper (mg / kg (ppm))	0572	16.300	2.0000	14.093	2.7236	0.71114	93	0.81	022.53	0
022	Copper (mg / kg (ppm))	0190	16.465	0.15000	14.093	2.7236	0.71114	93	0.87	022.42	0
022	Copper (mg / kg (ppm))	0083	16.500	1.0000	14.093	2.7236	0.71114	93	0.88	022.41	0
022	Copper (mg / kg (ppm))	0693	16.545	1.6500	14.093	2.7236	0.71114	93	0.90	022.42	0
022	Copper (mg / kg (ppm))	0278	16.550	0.10000	14.093	2.7236	0.71114	93	0.90	022.42	0
022	Copper (mg / kg (ppm))	0098	16.600	1.2000	14.093	2.7236	0.71114	93	0.92	022.43	0
022	Copper (mg / kg (ppm))	2058	16.830	1.8200	14.093	2.7236	0.71114	93	1.00	022.43	0
022	Copper (mg / kg (ppm))	0536	17.405	0.65000	14.093	2.7236	0.71114	93	1.22	022.31	0
022	Copper (mg / kg (ppm))	0964	17.480	1.4800	14.093	2.7236	0.71114	93	1.24	022.43	0
022	Copper (mg / kg (ppm))	0038	18.000	2.0000	14.093	2.7236	0.71114	93	1.43	022.32	0
022	Copper (mg / kg (ppm))	0035	18.030	1.0800	14.093	2.7236	0.71114	93	1.45	022.42	0
022	Copper (mg / kg (ppm))	0035	18.255	0.45000	14.093	2.7236	0.71114	93	1.53	022.32	0
022	Copper (mg / kg (ppm))	0187	19.420	0.14000	14.093	2.7236	0.71114	93	1.96	022.42	0
022	Copper (mg / kg (ppm))	0202	20.500	2.9600	14.093	2.7236	0.71114	93	2.35	022.42	0
022	Copper (mg / kg (ppm))	0169	20.500	0.40000	14.093	2.7236	0.71114	93	2.35	022.43	0
022	Copper (mg / kg (ppm))	2012	20.895	1.9100	14.093	2.7236	0.71114	93	2.50	022.41	0
022	Copper (mg / kg (ppm))	0106	21.275	0.65000	14.093	2.7236	0.71114	93	2.64	022.42	0
022	Copper (mg / kg (ppm))	0043	26.300	0.20000	14.093	2.7236	0.71114	93	4.48	022.43	0
022	Copper (mg / kg (ppm))	0042	30.400	2.4000	14.093	2.7236	0.71114	93	5.99	022.42	0
022	Copper (mg / kg (ppm))	0609	22.500	5.0000	14.093	2.7236	0.71114	93	3.09	022.32	1
023	Fluorine (%)	0619	0.00040	0.00000			0.00000	1			
025	Iron (mg / kg (ppm))	2023	143.00	0.00000	212.29	21.189	7.6604	95	-3.27	025.43	0
025	Iron (mg / kg (ppm))	0918	154.55	2.9600	212.29	21.189	7.6604	95	-2.73	025.43	0
025	Iron (mg / kg (ppm))	0921	169.95	18.950	212.29	21.189	7.6604	95	-2.00	025.31	0
025	Iron (mg / kg (ppm))	0154	172.75	33.390	212.29	21.189	7.6604	95	-1.87	025.52	0
025	Iron (mg / kg (ppm))	0563	175.50	9.0000	212.29	21.189	7.6604	95	-1.74	025.31	0
025	Iron (mg / kg (ppm))	0619	177.00	16.000	212.29	21.189	7.6604	95	-1.67	025.41	0
025	Iron (mg / kg (ppm))	0951	177.95	2.9000	212.29	21.189	7.6604	95	-1.62	025.31	0
025	Iron (mg / kg (ppm))	0629	180.50	0.00000	212.29	21.189	7.6604	95	-1.50	025.43	0
025	Iron (mg / kg (ppm))	0609	182.50	5.0000	212.29	21.189	7.6604	95	-1.41	025.32	0
025	Iron (mg / kg (ppm))	0265	183.00	6.0000	212.29	21.189	7.6604	95	-1.38	025.42	0
025	Iron (mg / kg (ppm))	0294	187.60	1.0000	212.29	21.189	7.6604	95	-1.17	025.42	0
025	Iron (mg / kg (ppm))	0345	188.00	4.0000	212.29	21.189	7.6604	95	-1.15	025.43	0



Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
025	Iron (mg / kg (ppm))	0297	189.50	7.0000	212.29	21.189	7.6604	95	-1.08	025.43	0
025	Iron (mg / kg (ppm))	0510	190.50	7.0000	212.29	21.189	7.6604	95	-1.03	025.43	0
025	Iron (mg / kg (ppm))	0505	191.50	3.0000	212.29	21.189	7.6604	95	-0.98	025.31	0
025	Iron (mg / kg (ppm))	0610	192.00	2.0000	212.29	21.189	7.6604	95	-0.96	025.41	0
025	Iron (mg / kg (ppm))	2012	193.14	1.2100	212.29	21.189	7.6604	95	-0.90	025.41	0
025	Iron (mg / kg (ppm))	0560	194.00	0.00000	212.29	21.189	7.6604	95	-0.86	025.42	0
025	Iron (mg / kg (ppm))	0187	194.28	1.3300	212.29	21.189	7.6604	95	-0.85	025.42	0
025	Iron (mg / kg (ppm))	0948	196.37	1.8200	212.29	21.189	7.6604	95	-0.75	025.33	0
025	Iron (mg / kg (ppm))	0555	196.75	2.0262	212.29	21.189	7.6604	95	-0.73	025.42	0
025	Iron (mg / kg (ppm))	0043	197.00	14.000	212.29	21.189	7.6604	95	-0.72	025.43	0
025	Iron (mg / kg (ppm))	0300	197.28	17.495	212.29	21.189	7.6604	95	-0.71	025.41	0
025	Iron (mg / kg (ppm))	0726	197.40	0.60000	212.29	21.189	7.6604	95	-0.70	025.42	0
025	Iron (mg / kg (ppm))	0504	198.40	30.600	212.29	21.189	7.6604	95	-0.66	025.33	0
025	Iron (mg / kg (ppm))	0638	199.50	5.0000	212.29	21.189	7.6604	95	-0.60	025.31	0
025	Iron (mg / kg (ppm))	0720	199.65	7.9000	212.29	21.189	7.6604	95	-0.60	025.41	0
025	Iron (mg / kg (ppm))	0629	200.00	0.00000	212.29	21.189	7.6604	95	-0.58	025.31	0
025	Iron (mg / kg (ppm))	0874	201.00	6.0000	212.29	21.189	7.6604	95	-0.53	025.31	0
025	Iron (mg / kg (ppm))	0413	202.50	13.000	212.29	21.189	7.6604	95	-0.46	025.42	0
025	Iron (mg / kg (ppm))	0598	202.95	8.9000	212.29	21.189	7.6604	95	-0.44	025.41	0
025	Iron (mg / kg (ppm))	0160	203.00	12.000	212.29	21.189	7.6604	95	-0.44	025.42	0
025	Iron (mg / kg (ppm))	0164	203.50	7.0000	212.29	21.189	7.6604	95	-0.41	025.41	0
025	Iron (mg / kg (ppm))	2058	203.87	8.0900	212.29	21.189	7.6604	95	-0.40	025.43	0
025	Iron (mg / kg (ppm))	0695	203.90	2.0300	212.29	21.189	7.6604	95	-0.40	025.41	0
025	Iron (mg / kg (ppm))	0692	205.50	1.0000	212.29	21.189	7.6604	95	-0.32	025.42	0
025	Iron (mg / kg (ppm))	0682	205.87	3.9000	212.29	21.189	7.6604	95	-0.30	025.41	0
025	Iron (mg / kg (ppm))	0011	206.21	2.1300	212.29	21.189	7.6604	95	-0.29	025.41	0
025	Iron (mg / kg (ppm))	0190	206.30	2.8000	212.29	21.189	7.6604	95	-0.28	025.42	0
025	Iron (mg / kg (ppm))	0242	206.50	5.0000	212.29	21.189	7.6604	95	-0.27	025.99	0
025	Iron (mg / kg (ppm))	0098	207.75	16.700	212.29	21.189	7.6604	95	-0.21	025.43	0
025	Iron (mg / kg (ppm))	0407	208.74	5.2400	212.29	21.189	7.6604	95	-0.17	025.41	0
025	Iron (mg / kg (ppm))	0511	210.50	3.0000	212.29	21.189	7.6604	95	-0.08	025.41	0
025	Iron (mg / kg (ppm))	0229	211.00	2.0000	212.29	21.189	7.6604	95	-0.06	025.41	0
025	Iron (mg / kg (ppm))	0358	211.05	0.29000	212.29	21.189	7.6604	95	-0.06	025.41	0
025	Iron (mg / kg (ppm))	0098	211.80	6.6000	212.29	21.189	7.6604	95	-0.02	025.41	0
025	Iron (mg / kg (ppm))	0968	212.50	1.0000	212.29	21.189	7.6604	95	0.01	025.43	0
025	Iron (mg / kg (ppm))	0051	213.00	2.0000	212.29	21.189	7.6604	95	0.03	025.41	0
025	Iron (mg / kg (ppm))	0226	213.50	3.0000	212.29	21.189	7.6604	95	0.06	025.41	0
025	Iron (mg / kg (ppm))	0083	214.00	6.0000	212.29	21.189	7.6604	95	0.08	025.41	0
025	Iron (mg / kg (ppm))	0278	214.50	3.0000	212.29	21.189	7.6604	95	0.10	025.42	0
025	Iron (mg / kg (ppm))	0868	215.00	0.00000	212.29	21.189	7.6604	95	0.13	025.31	0
025	Iron (mg / kg (ppm))	0689	215.55	3.7000	212.29	21.189	7.6604	95	0.15	025.31	0
025	Iron (mg / kg (ppm))	0910	216.00	14.000	212.29	21.189	7.6604	95	0.18	025.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
025	Iron (mg / kg (ppm))	0553	216.50	1.0000	212.29	21.189	7.6604	95	0.20	025.41	0
025	Iron (mg / kg (ppm))	2004	216.50	7.0000	212.29	21.189	7.6604	95	0.20	025.41	0
025	Iron (mg / kg (ppm))	0004	217.00	6.0000	212.29	21.189	7.6604	95	0.22	025.41	0
025	Iron (mg / kg (ppm))	0693	217.00	4.0000	212.29	21.189	7.6604	95	0.22	025.42	0
025	Iron (mg / kg (ppm))	0074	218.00	2.0000	212.29	21.189	7.6604	95	0.27	025.41	0
025	Iron (mg / kg (ppm))	0354	219.49	0.36000	212.29	21.189	7.6604	95	0.34	025.31	0
025	Iron (mg / kg (ppm))	0049	219.92	4.4700	212.29	21.189	7.6604	95	0.36	025.41	0
025	Iron (mg / kg (ppm))	0035	220.35	12.500	212.29	21.189	7.6604	95	0.38	025.42	0
025	Iron (mg / kg (ppm))	0353	220.92	15.430	212.29	21.189	7.6604	95	0.41	025.43	0
025	Iron (mg / kg (ppm))	0100	223.00	4.0000	212.29	21.189	7.6604	95	0.51	025.41	0
025	Iron (mg / kg (ppm))	0171	223.00	22.000	212.29	21.189	7.6604	95	0.51	025.41	0
025	Iron (mg / kg (ppm))	0033	223.00	6.0000	212.29	21.189	7.6604	95	0.51	025.43	0
025	Iron (mg / kg (ppm))	0178	223.90	3.2000	212.29	21.189	7.6604	95	0.55	025.31	0
025	Iron (mg / kg (ppm))	0008	224.50	5.0000	212.29	21.189	7.6604	95	0.58	025.43	0
025	Iron (mg / kg (ppm))	0035	224.95	15.300	212.29	21.189	7.6604	95	0.60	025.32	0
025	Iron (mg / kg (ppm))	0878	225.00	10.000	212.29	21.189	7.6604	95	0.60	025.41	0
025	Iron (mg / kg (ppm))	0148	225.10	0.20000	212.29	21.189	7.6604	95	0.60	025.41	0
025	Iron (mg / kg (ppm))	0653	225.30	1.4000	212.29	21.189	7.6604	95	0.61	025.31	0
025	Iron (mg / kg (ppm))	0208	229.25	2.1000	212.29	21.189	7.6604	95	0.80	025.41	0
025	Iron (mg / kg (ppm))	0939	229.26	2.5100	212.29	21.189	7.6604	95	0.80	025.31	0
025	Iron (mg / kg (ppm))	0106	229.34	26.670	212.29	21.189	7.6604	95	0.80	025.42	0
025	Iron (mg / kg (ppm))	0208	230.00	2.0000	212.29	21.189	7.6604	95	0.84	025.31	0
025	Iron (mg / kg (ppm))	0504	230.85	13.900	212.29	21.189	7.6604	95	0.88	025.32	0
025	Iron (mg / kg (ppm))	0038	231.00	10.000	212.29	21.189	7.6604	95	0.88	025.43	0
025	Iron (mg / kg (ppm))	0425	231.05	9.9000	212.29	21.189	7.6604	95	0.89	025.43	0
025	Iron (mg / kg (ppm))	0038	232.50	7.0000	212.29	21.189	7.6604	95	0.95	025.31	0
025	Iron (mg / kg (ppm))	0964	233.08	8.1100	212.29	21.189	7.6604	95	0.98	025.41	0
025	Iron (mg / kg (ppm))	0199	234.00	16.000	212.29	21.189	7.6604	95	1.02	025.53	0
025	Iron (mg / kg (ppm))	0520	234.50	11.000	212.29	21.189	7.6604	95	1.05	025.41	0
025	Iron (mg / kg (ppm))	0337	235.12	0.72000	212.29	21.189	7.6604	95	1.08	025.32	0
025	Iron (mg / kg (ppm))	0045	236.00	18.000	212.29	21.189	7.6604	95	1.12	025.42	0
025	Iron (mg / kg (ppm))	0168	240.00	40.000	212.29	21.189	7.6604	95	1.31	025.43	0
025	Iron (mg / kg (ppm))	2018	240.01	4.0300	212.29	21.189	7.6604	95	1.31	025.31	0
025	Iron (mg / kg (ppm))	0964	242.32	29.110	212.29	21.189	7.6604	95	1.42	025.43	0
025	Iron (mg / kg (ppm))	0003	243.00	2.0000	212.29	21.189	7.6604	95	1.45	025.41	0
025	Iron (mg / kg (ppm))	0675	254.36	24.180	212.29	21.189	7.6604	95	1.99	025.43	0
025	Iron (mg / kg (ppm))	0169	263.50	9.0000	212.29	21.189	7.6604	95	2.42	025.43	0
025	Iron (mg / kg (ppm))	0066	263.97	1.5300	212.29	21.189	7.6604	95	2.44	025.31	0
025	Iron (mg / kg (ppm))	0405	274.50	9.0000	212.29	21.189	7.6604	95	2.94	025.41	0
025	Iron (mg / kg (ppm))	0175	285.00	10.000	212.29	21.189	7.6604	95	3.43	025.31	0
025	Iron (mg / kg (ppm))	2067	340.20	12.554	212.29	21.189	7.6604	95	6.04	025.31	0
025	Iron (mg / kg (ppm))	2055	126.55	50.300	212.29	21.189	7.6604	95	-4.05	025.99	1

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
025	Iron (mg / kg (ppm))	2022	234.50	43.000	212.29	21.189	7.6604	95	1.05	025.31	1
025	Iron (mg / kg (ppm))	0042	248.50	65.000	212.29	21.189	7.6604	95	1.71	025.42	1
027	Magnesium (%)	0510	0.29500	0.01000	0.36563	0.01694	0.00889	88	-4.17	027.43	0
027	Magnesium (%)	0043	0.32000	0.00000	0.36563	0.01694	0.00889	88	-2.69	027.43	0
027	Magnesium (%)	0154	0.32390	0.00340	0.36563	0.01694	0.00889	88	-2.46	027.52	0
027	Magnesium (%)	2055	0.32900	0.00400	0.36563	0.01694	0.00889	88	-2.16	027.99	0
027	Magnesium (%)	0921	0.33150	0.00700	0.36563	0.01694	0.00889	88	-2.01	027.34	0
027	Magnesium (%)	0675	0.33500	0.01000	0.36563	0.01694	0.00889	88	-1.81	027.43	0
027	Magnesium (%)	0345	0.33750	0.00300	0.36563	0.01694	0.00889	88	-1.66	027.43	0
027	Magnesium (%)	0918	0.33950	0.01300	0.36563	0.01694	0.00889	88	-1.54	027.43	0
027	Magnesium (%)	0695	0.34000	0.00000	0.36563	0.01694	0.00889	88	-1.51	027.41	0
027	Magnesium (%)	0563	0.34125	0.00150	0.36563	0.01694	0.00889	88	-1.44	027.31	0
027	Magnesium (%)	0065	0.34265	0.01070	0.36563	0.01694	0.00889	88	-1.36	027.31	0
027	Magnesium (%)	0358	0.34500	0.01000	0.36563	0.01694	0.00889	88	-1.22	027.41	0
027	Magnesium (%)	0598	0.34500	0.01120	0.36563	0.01694	0.00889	88	-1.22	027.41	0
027	Magnesium (%)	0692	0.34500	0.01000	0.36563	0.01694	0.00889	88	-1.22	027.42	0
027	Magnesium (%)	0553	0.34900	0.01400	0.36563	0.01694	0.00889	88	-0.98	027.41	0
027	Magnesium (%)	0038	0.35000	0.00400	0.36563	0.01694	0.00889	88	-0.92	027.31	0
027	Magnesium (%)	2058	0.35000	0.02000	0.36563	0.01694	0.00889	88	-0.92	027.43	0
027	Magnesium (%)	0038	0.35100	0.01800	0.36563	0.01694	0.00889	88	-0.86	027.43	0
027	Magnesium (%)	0142	0.35200	0.00200	0.36563	0.01694	0.00889	88	-0.80	027.31	0
027	Magnesium (%)	0035	0.35245	0.00730	0.36563	0.01694	0.00889	88	-0.78	027.32	0
027	Magnesium (%)	2004	0.35450	0.01100	0.36563	0.01694	0.00889	88	-0.66	027.44	0
027	Magnesium (%)	0175	0.35500	0.01000	0.36563	0.01694	0.00889	88	-0.63	027.31	0
027	Magnesium (%)	0353	0.35500	0.03000	0.36563	0.01694	0.00889	88	-0.63	027.43	0
027	Magnesium (%)	0242	0.35500	0.01000	0.36563	0.01694	0.00889	88	-0.63	027.99	0
027	Magnesium (%)	0868	0.35600	0.00000	0.36563	0.01694	0.00889	88	-0.57	027.31	0
027	Magnesium (%)	0504	0.35610	0.02260	0.36563	0.01694	0.00889	88	-0.56	027.33	0
027	Magnesium (%)	0045	0.35800	0.01400	0.36563	0.01694	0.00889	88	-0.45	027.42	0
027	Magnesium (%)	0354	0.36000	0.00000	0.36563	0.01694	0.00889	88	-0.33	027.31	0
027	Magnesium (%)	0638	0.36000	0.00000	0.36563	0.01694	0.00889	88	-0.33	027.31	0
027	Magnesium (%)	0689	0.36000	0.00000	0.36563	0.01694	0.00889	88	-0.33	027.31	0
027	Magnesium (%)	0049	0.36000	0.00000	0.36563	0.01694	0.00889	88	-0.33	027.41	0
027	Magnesium (%)	0148	0.36000	0.00000	0.36563	0.01694	0.00889	88	-0.33	027.41	0
027	Magnesium (%)	0357	0.36000	0.00000	0.36563	0.01694	0.00889	88	-0.33	027.42	0
027	Magnesium (%)	0297	0.36000	0.00000	0.36563	0.01694	0.00889	88	-0.33	027.43	0
027	Magnesium (%)	0035	0.36020	0.00060	0.36563	0.01694	0.00889	88	-0.32	027.42	0
027	Magnesium (%)	0407	0.36050	0.00700	0.36563	0.01694	0.00889	88	-0.30	027.41	0
027	Magnesium (%)	0300	0.36100	0.00800	0.36563	0.01694	0.00889	88	-0.27	027.41	0
027	Magnesium (%)	0164	0.36200	0.00400	0.36563	0.01694	0.00889	88	-0.21	027.41	0
027	Magnesium (%)	0008	0.36200	0.01200	0.36563	0.01694	0.00889	88	-0.21	027.43	0
027	Magnesium (%)	0042	0.36300	0.02800	0.36563	0.01694	0.00889	88	-0.16	027.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
027	Magnesium (%)	0011	0.36465	0.00930	0.36563	0.01694	0.00889	88	-0.06	027.41	0
027	Magnesium (%)	0536	0.36500	0.03000	0.36563	0.01694	0.00889	88	-0.04	027.31	0
027	Magnesium (%)	2018	0.36500	0.01000	0.36563	0.01694	0.00889	88	-0.04	027.31	0
027	Magnesium (%)	0609	0.36500	0.01000	0.36563	0.01694	0.00889	88	-0.04	027.32	0
027	Magnesium (%)	0171	0.36500	0.03000	0.36563	0.01694	0.00889	88	-0.04	027.41	0
027	Magnesium (%)	0229	0.36500	0.01000	0.36563	0.01694	0.00889	88	-0.04	027.41	0
027	Magnesium (%)	0878	0.36500	0.01000	0.36563	0.01694	0.00889	88	-0.04	027.41	0
027	Magnesium (%)	0278	0.36500	0.01000	0.36563	0.01694	0.00889	88	-0.04	027.42	0
027	Magnesium (%)	0619	0.36550	0.00300	0.36563	0.01694	0.00889	88	-0.01	027.41	0
027	Magnesium (%)	0139	0.36715	0.00730	0.36563	0.01694	0.00889	88	0.09	027.31	0
027	Magnesium (%)	0650	0.36725	0.01790	0.36563	0.01694	0.00889	88	0.10	027.31	0
027	Magnesium (%)	0653	0.36800	0.00400	0.36563	0.01694	0.00889	88	0.14	027.31	0
027	Magnesium (%)	0208	0.36950	0.00300	0.36563	0.01694	0.00889	88	0.23	027.31	0
027	Magnesium (%)	0208	0.36950	0.00500	0.36563	0.01694	0.00889	88	0.23	027.41	0
027	Magnesium (%)	0178	0.37000	0.00000	0.36563	0.01694	0.00889	88	0.26	027.31	0
027	Magnesium (%)	0098	0.37000	0.00000	0.36563	0.01694	0.00889	88	0.26	027.41	0
027	Magnesium (%)	0720	0.37000	0.00000	0.36563	0.01694	0.00889	88	0.26	027.41	0
027	Magnesium (%)	2012	0.37000	0.00000	0.36563	0.01694	0.00889	88	0.26	027.41	0
027	Magnesium (%)	0294	0.37000	0.00000	0.36563	0.01694	0.00889	88	0.26	027.42	0
027	Magnesium (%)	0037	0.37000	0.00000	0.36563	0.01694	0.00889	88	0.26	027.43	0
027	Magnesium (%)	0033	0.37100	0.00400	0.36563	0.01694	0.00889	88	0.32	027.43	0
027	Magnesium (%)	0560	0.37285	0.00630	0.36563	0.01694	0.00889	88	0.43	027.42	0
027	Magnesium (%)	2023	0.37350	0.00900	0.36563	0.01694	0.00889	88	0.46	027.44	0
027	Magnesium (%)	0083	0.37500	0.01000	0.36563	0.01694	0.00889	88	0.55	027.41	0
027	Magnesium (%)	0693	0.37500	0.01000	0.36563	0.01694	0.00889	88	0.55	027.42	0
027	Magnesium (%)	0629	0.37500	0.01000	0.36563	0.01694	0.00889	88	0.55	027.43	0
027	Magnesium (%)	0263	0.38140	0.00140	0.36563	0.01694	0.00889	88	0.93	027.32	0
027	Magnesium (%)	0572	0.38175	0.00430	0.36563	0.01694	0.00889	88	0.95	027.53	0
027	Magnesium (%)	0003	0.38500	0.03000	0.36563	0.01694	0.00889	88	1.14	027.41	0
027	Magnesium (%)	0098	0.38500	0.03000	0.36563	0.01694	0.00889	88	1.14	027.43	0
027	Magnesium (%)	0425	0.38500	0.01000	0.36563	0.01694	0.00889	88	1.14	027.43	0
027	Magnesium (%)	0964	0.38550	0.00180	0.36563	0.01694	0.00889	88	1.17	027.41	0
027	Magnesium (%)	0187	0.38590	0.00220	0.36563	0.01694	0.00889	88	1.20	027.42	0
027	Magnesium (%)	0555	0.38820	0.00360	0.36563	0.01694	0.00889	88	1.33	027.42	0
027	Magnesium (%)	0968	0.38850	0.00100	0.36563	0.01694	0.00889	88	1.35	027.43	0
027	Magnesium (%)	0106	0.39000	0.00000	0.36563	0.01694	0.00889	88	1.44	027.42	0
027	Magnesium (%)	0889	0.39000	0.02000	0.36563	0.01694	0.00889	88	1.44	027.99	0
027	Magnesium (%)	0964	0.39055	0.03450	0.36563	0.01694	0.00889	88	1.47	027.43	0
027	Magnesium (%)	0504	0.39070	0.02560	0.36563	0.01694	0.00889	88	1.48	027.32	0
027	Magnesium (%)	0226	0.39500	0.01000	0.36563	0.01694	0.00889	88	1.73	027.41	0
027	Magnesium (%)	0413	0.39500	0.01000	0.36563	0.01694	0.00889	88	1.73	027.41	0
027	Magnesium (%)	0202	0.39500	0.01000	0.36563	0.01694	0.00889	88	1.73	027.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
027	Magnesium (%)	0100	0.40000	0.00000	0.36563	0.01694	0.00889	88	2.03	027.41	0
027	Magnesium (%)	0682	0.40000	0.00000	0.36563	0.01694	0.00889	88	2.03	027.41	0
027	Magnesium (%)	0405	0.40500	0.01000	0.36563	0.01694	0.00889	88	2.32	027.41	0
027	Magnesium (%)	0874	0.41480	0.00300	0.36563	0.01694	0.00889	88	2.90	027.31	0
027	Magnesium (%)	0520	0.41500	0.03000	0.36563	0.01694	0.00889	88	2.91	027.41	0
027	Magnesium (%)	0948	0.52000	0.02000	0.36563	0.01694	0.00889	88	9.11	027.33	0
027	Magnesium (%)	0265	0.37000	0.04000	0.36563	0.01694	0.00889	88	0.26	027.42	1
027	Magnesium (%)	0051	1.0948	0.00120	0.36563	0.01694	0.00889	88	43.04	027.41	2
028	Manganese (mg / kg (ppm))	2022	103.50	1.0000	124.03	8.7096	2.5869	97	-2.36	028.31	0
028	Manganese (mg / kg (ppm))	0951	105.50	0.00000	124.03	8.7096	2.5869	97	-2.13	028.31	0
028	Manganese (mg / kg (ppm))	0695	106.39	0.79000	124.03	8.7096	2.5869	97	-2.03	028.41	0
028	Manganese (mg / kg (ppm))	0511	110.00	8.0000	124.03	8.7096	2.5869	97	-1.61	028.41	0
028	Manganese (mg / kg (ppm))	0921	111.02	2.0300	124.03	8.7096	2.5869	97	-1.49	028.31	0
028	Manganese (mg / kg (ppm))	0939	113.23	2.6000	124.03	8.7096	2.5869	97	-1.24	028.31	0
028	Manganese (mg / kg (ppm))	0178	113.35	0.90000	124.03	8.7096	2.5869	97	-1.23	028.31	0
028	Manganese (mg / kg (ppm))	0553	113.50	1.0000	124.03	8.7096	2.5869	97	-1.21	028.41	0
028	Manganese (mg / kg (ppm))	0294	113.58	2.5600	124.03	8.7096	2.5869	97	-1.20	028.42	0
028	Manganese (mg / kg (ppm))	2058	114.09	0.21000	124.03	8.7096	2.5869	97	-1.14	028.43	0
028	Manganese (mg / kg (ppm))	0049	114.13	2.6300	124.03	8.7096	2.5869	97	-1.14	028.41	0
028	Manganese (mg / kg (ppm))	0720	114.92	0.95000	124.03	8.7096	2.5869	97	-1.05	028.41	0
028	Manganese (mg / kg (ppm))	0164	115.00	2.0000	124.03	8.7096	2.5869	97	-1.04	028.41	0
028	Manganese (mg / kg (ppm))	0168	115.00	10.000	124.03	8.7096	2.5869	97	-1.04	028.43	0
028	Manganese (mg / kg (ppm))	0510	115.00	2.0000	124.03	8.7096	2.5869	97	-1.04	028.43	0
028	Manganese (mg / kg (ppm))	0563	115.50	1.0000	124.03	8.7096	2.5869	97	-0.98	028.31	0
028	Manganese (mg / kg (ppm))	0629	115.50	1.0000	124.03	8.7096	2.5869	97	-0.98	028.31	0
028	Manganese (mg / kg (ppm))	0358	115.81	5.6900	124.03	8.7096	2.5869	97	-0.94	028.41	0
028	Manganese (mg / kg (ppm))	0407	115.93	0.66000	124.03	8.7096	2.5869	97	-0.93	028.41	0
028	Manganese (mg / kg (ppm))	0043	116.50	1.0000	124.03	8.7096	2.5869	97	-0.87	028.43	0
028	Manganese (mg / kg (ppm))	0629	117.00	0.00000	124.03	8.7096	2.5869	97	-0.81	028.43	0
028	Manganese (mg / kg (ppm))	0208	117.30	2.4000	124.03	8.7096	2.5869	97	-0.77	028.41	0
028	Manganese (mg / kg (ppm))	2012	117.38	2.4700	124.03	8.7096	2.5869	97	-0.76	028.41	0
028	Manganese (mg / kg (ppm))	0590	117.50	1.0000	124.03	8.7096	2.5869	97	-0.75	028.31	0
028	Manganese (mg / kg (ppm))	0300	117.92	4.6410	124.03	8.7096	2.5869	97	-0.70	028.41	0
028	Manganese (mg / kg (ppm))	2055	117.95	4.9000	124.03	8.7096	2.5869	97	-0.70	028.99	0
028	Manganese (mg / kg (ppm))	0638	118.50	1.0000	124.03	8.7096	2.5869	97	-0.64	028.31	0
028	Manganese (mg / kg (ppm))	0610	118.50	1.0000	124.03	8.7096	2.5869	97	-0.64	028.41	0
028	Manganese (mg / kg (ppm))	2023	118.50	3.0000	124.03	8.7096	2.5869	97	-0.64	028.44	0
028	Manganese (mg / kg (ppm))	0098	118.55	2.1000	124.03	8.7096	2.5869	97	-0.63	028.41	0
028	Manganese (mg / kg (ppm))	0689	118.75	2.1000	124.03	8.7096	2.5869	97	-0.61	028.31	0
028	Manganese (mg / kg (ppm))	0051	118.82	0.68000	124.03	8.7096	2.5869	97	-0.60	028.41	0
028	Manganese (mg / kg (ppm))	2018	119.16	3.6100	124.03	8.7096	2.5869	97	-0.56	028.31	0
028	Manganese (mg / kg (ppm))	0555	119.99	3.0620	124.03	8.7096	2.5869	97	-0.46	028.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
028	Manganese (mg / kg (ppm))	0354	120.00	0.31000	124.03	8.7096	2.5869	97	-0.46	028.31	0
028	Manganese (mg / kg (ppm))	0609	120.00	10.000	124.03	8.7096	2.5869	97	-0.46	028.32	0
028	Manganese (mg / kg (ppm))	0598	120.05	5.1000	124.03	8.7096	2.5869	97	-0.46	028.41	0
028	Manganese (mg / kg (ppm))	0505	121.00	2.0000	124.03	8.7096	2.5869	97	-0.35	028.31	0
028	Manganese (mg / kg (ppm))	0074	121.00	2.0000	124.03	8.7096	2.5869	97	-0.35	028.41	0
028	Manganese (mg / kg (ppm))	0619	121.50	5.0000	124.03	8.7096	2.5869	97	-0.29	028.41	0
028	Manganese (mg / kg (ppm))	0910	121.50	3.0000	124.03	8.7096	2.5869	97	-0.29	028.41	0
028	Manganese (mg / kg (ppm))	0148	121.60	0.60000	124.03	8.7096	2.5869	97	-0.28	028.41	0
028	Manganese (mg / kg (ppm))	0948	121.67	0.93000	124.03	8.7096	2.5869	97	-0.27	028.33	0
028	Manganese (mg / kg (ppm))	0693	122.50	5.0000	124.03	8.7096	2.5869	97	-0.18	028.42	0
028	Manganese (mg / kg (ppm))	0918	122.79	2.2500	124.03	8.7096	2.5869	97	-0.14	028.43	0
028	Manganese (mg / kg (ppm))	0011	122.88	0.74500	124.03	8.7096	2.5869	97	-0.13	028.41	0
028	Manganese (mg / kg (ppm))	2004	123.00	2.0000	124.03	8.7096	2.5869	97	-0.12	028.44	0
028	Manganese (mg / kg (ppm))	0004	123.50	1.0000	124.03	8.7096	2.5869	97	-0.06	028.41	0
028	Manganese (mg / kg (ppm))	2067	123.85	5.0790	124.03	8.7096	2.5869	97	-0.02	028.31	0
028	Manganese (mg / kg (ppm))	0229	124.00	4.0000	124.03	8.7096	2.5869	97	0.00	028.41	0
028	Manganese (mg / kg (ppm))	0723	124.50	0.00000	124.03	8.7096	2.5869	97	0.05	028.31	0
028	Manganese (mg / kg (ppm))	0003	124.50	3.0000	124.03	8.7096	2.5869	97	0.05	028.41	0
028	Manganese (mg / kg (ppm))	0066	124.88	2.3900	124.03	8.7096	2.5869	97	0.10	028.31	0
028	Manganese (mg / kg (ppm))	0692	125.00	6.0000	124.03	8.7096	2.5869	97	0.11	028.42	0
028	Manganese (mg / kg (ppm))	0675	125.24	0.21000	124.03	8.7096	2.5869	97	0.14	028.43	0
028	Manganese (mg / kg (ppm))	0353	125.50	1.4000	124.03	8.7096	2.5869	97	0.17	028.43	0
028	Manganese (mg / kg (ppm))	0504	125.95	1.5000	124.03	8.7096	2.5869	97	0.22	028.33	0
028	Manganese (mg / kg (ppm))	0171	126.50	7.0000	124.03	8.7096	2.5869	97	0.28	028.41	0
028	Manganese (mg / kg (ppm))	0413	126.50	7.0000	124.03	8.7096	2.5869	97	0.28	028.42	0
028	Manganese (mg / kg (ppm))	0968	126.50	1.0000	124.03	8.7096	2.5869	97	0.28	028.43	0
028	Manganese (mg / kg (ppm))	0265	127.00	4.0000	124.03	8.7096	2.5869	97	0.34	028.42	0
028	Manganese (mg / kg (ppm))	0868	127.50	1.0000	124.03	8.7096	2.5869	97	0.40	028.31	0
028	Manganese (mg / kg (ppm))	0297	127.50	1.0000	124.03	8.7096	2.5869	97	0.40	028.43	0
028	Manganese (mg / kg (ppm))	0278	128.00	2.0000	124.03	8.7096	2.5869	97	0.46	028.42	0
028	Manganese (mg / kg (ppm))	0964	128.07	1.0300	124.03	8.7096	2.5869	97	0.46	028.41	0
028	Manganese (mg / kg (ppm))	0242	128.50	5.0000	124.03	8.7096	2.5869	97	0.51	028.99	0
028	Manganese (mg / kg (ppm))	0038	129.50	1.0000	124.03	8.7096	2.5869	97	0.63	028.32	0
028	Manganese (mg / kg (ppm))	0520	129.50	5.0000	124.03	8.7096	2.5869	97	0.63	028.41	0
028	Manganese (mg / kg (ppm))	0035	129.95	2.7000	124.03	8.7096	2.5869	97	0.68	028.42	0
028	Manganese (mg / kg (ppm))	0083	130.00	4.0000	124.03	8.7096	2.5869	97	0.68	028.41	0
028	Manganese (mg / kg (ppm))	0878	130.00	0.00000	124.03	8.7096	2.5869	97	0.68	028.41	0
028	Manganese (mg / kg (ppm))	0008	130.05	2.9000	124.03	8.7096	2.5869	97	0.69	028.43	0
028	Manganese (mg / kg (ppm))	0425	130.40	0.80000	124.03	8.7096	2.5869	97	0.73	028.43	0
028	Manganese (mg / kg (ppm))	0208	130.50	3.0000	124.03	8.7096	2.5869	97	0.74	028.31	0
028	Manganese (mg / kg (ppm))	0357	130.50	1.0000	124.03	8.7096	2.5869	97	0.74	028.42	0
028	Manganese (mg / kg (ppm))	0035	130.95	3.3000	124.03	8.7096	2.5869	97	0.79	028.32	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
028	Manganese (mg / kg (ppm))	0042	131.50	3.0000	124.03	8.7096	2.5869	97	0.86	028.42	0
028	Manganese (mg / kg (ppm))	0033	131.50	3.0000	124.03	8.7096	2.5869	97	0.86	028.43	0
028	Manganese (mg / kg (ppm))	0038	131.50	3.0000	124.03	8.7096	2.5869	97	0.86	028.43	0
028	Manganese (mg / kg (ppm))	0345	131.50	3.0000	124.03	8.7096	2.5869	97	0.86	028.43	0
028	Manganese (mg / kg (ppm))	0199	131.80	1.8000	124.03	8.7096	2.5869	97	0.89	028.53	0
028	Manganese (mg / kg (ppm))	0504	131.95	3.9000	124.03	8.7096	2.5869	97	0.91	028.32	0
028	Manganese (mg / kg (ppm))	0226	132.50	7.0000	124.03	8.7096	2.5869	97	0.97	028.41	0
028	Manganese (mg / kg (ppm))	0572	133.00	8.0000	124.03	8.7096	2.5869	97	1.03	028.53	0
028	Manganese (mg / kg (ppm))	0175	133.50	3.0000	124.03	8.7096	2.5869	97	1.09	028.31	0
028	Manganese (mg / kg (ppm))	0682	133.95	2.7800	124.03	8.7096	2.5869	97	1.14	028.41	0
028	Manganese (mg / kg (ppm))	0190	135.26	1.6900	124.03	8.7096	2.5869	97	1.29	028.42	0
028	Manganese (mg / kg (ppm))	0045	135.50	1.0000	124.03	8.7096	2.5869	97	1.32	028.42	0
028	Manganese (mg / kg (ppm))	0536	135.75	1.3000	124.03	8.7096	2.5869	97	1.35	028.31	0
028	Manganese (mg / kg (ppm))	0202	137.00	2.0200	124.03	8.7096	2.5869	97	1.49	028.42	0
028	Manganese (mg / kg (ppm))	0560	137.00	2.0000	124.03	8.7096	2.5869	97	1.49	028.42	0
028	Manganese (mg / kg (ppm))	0726	137.85	0.10000	124.03	8.7096	2.5869	97	1.59	028.42	0
028	Manganese (mg / kg (ppm))	0187	138.04	2.6100	124.03	8.7096	2.5869	97	1.61	028.42	0
028	Manganese (mg / kg (ppm))	0100	138.50	1.0000	124.03	8.7096	2.5869	97	1.66	028.41	0
028	Manganese (mg / kg (ppm))	0405	141.00	2.0000	124.03	8.7096	2.5869	97	1.95	028.41	0
028	Manganese (mg / kg (ppm))	0106	146.25	0.50000	124.03	8.7096	2.5869	97	2.55	028.42	0
028	Manganese (mg / kg (ppm))	0169	175.00	4.0000	124.03	8.7096	2.5869	97	5.85	028.43	0
028	Manganese (mg / kg (ppm))	0098	132.90	11.400	124.03	8.7096	2.5869	97	1.02	028.43	1
028	Manganese (mg / kg (ppm))	0964	141.31	13.720	124.03	8.7096	2.5869	97	1.98	028.43	1
031	Phosphorus (%)	0852	0.63500	0.01000	0.81719	0.03733	0.01604	157	-4.88	031.99	0
031	Phosphorus (%)	0692	0.67500	0.05000	0.81719	0.03733	0.01604	157	-3.81	031.42	0
031	Phosphorus (%)	0242	0.67500	0.05000	0.81719	0.03733	0.01604	157	-3.81	031.99	0
031	Phosphorus (%)	0108	0.68500	0.01000	0.81719	0.03733	0.01604	157	-3.54	031.01	0
031	Phosphorus (%)	0089	0.70000	0.00000	0.81719	0.03733	0.01604	157	-3.14	031.41	0
031	Phosphorus (%)	0646	0.72500	0.01000	0.81719	0.03733	0.01604	157	-2.47	031.01	0
031	Phosphorus (%)	0848	0.73500	0.01000	0.81719	0.03733	0.01604	157	-2.20	031.41	0
031	Phosphorus (%)	0874	0.74415	0.02590	0.81719	0.03733	0.01604	157	-1.96	031.01	0
031	Phosphorus (%)	0685	0.74500	0.01000	0.81719	0.03733	0.01604	157	-1.93	031.41	0
031	Phosphorus (%)	0043	0.74500	0.01000	0.81719	0.03733	0.01604	157	-1.93	031.43	0
031	Phosphorus (%)	0154	0.75000	0.01600	0.81719	0.03733	0.01604	157	-1.80	031.52	0
031	Phosphorus (%)	0510	0.75500	0.03000	0.81719	0.03733	0.01604	157	-1.67	031.43	0
031	Phosphorus (%)	0910	0.76000	0.08000	0.81719	0.03733	0.01604	157	-1.53	031.41	0
031	Phosphorus (%)	0693	0.76500	0.01000	0.81719	0.03733	0.01604	157	-1.40	031.42	0
031	Phosphorus (%)	2058	0.76500	0.03000	0.81719	0.03733	0.01604	157	-1.40	031.43	0
031	Phosphorus (%)	0139	0.76900	0.04200	0.81719	0.03733	0.01604	157	-1.29	031.01	0
031	Phosphorus (%)	0885	0.77000	0.02000	0.81719	0.03733	0.01604	157	-1.26	031.01	0
031	Phosphorus (%)	2055	0.77100	0.02400	0.81719	0.03733	0.01604	157	-1.24	031.99	0
031	Phosphorus (%)	2018	0.77500	0.03000	0.81719	0.03733	0.01604	157	-1.13	031.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
031	Phosphorus (%)	0144	0.77500	0.01000	0.81719	0.03733	0.01604	157	-1.13	031.41	0
031	Phosphorus (%)	0294	0.77500	0.01000	0.81719	0.03733	0.01604	157	-1.13	031.42	0
031	Phosphorus (%)	0017	0.77500	0.01000	0.81719	0.03733	0.01604	157	-1.13	031.43	0
031	Phosphorus (%)	0035	0.77550	0.02900	0.81719	0.03733	0.01604	157	-1.12	031.01	0
031	Phosphorus (%)	0042	0.77650	0.02700	0.81719	0.03733	0.01604	157	-1.09	031.42	0
031	Phosphorus (%)	0511	0.78000	0.02000	0.81719	0.03733	0.01604	157	-1.00	031.01	0
031	Phosphorus (%)	0638	0.78000	0.00000	0.81719	0.03733	0.01604	157	-1.00	031.01	0
031	Phosphorus (%)	0650	0.78000	0.02000	0.81719	0.03733	0.01604	157	-1.00	031.01	0
031	Phosphorus (%)	0169	0.78000	0.00000	0.81719	0.03733	0.01604	157	-1.00	031.03	0
031	Phosphorus (%)	0298	0.78000	0.02000	0.81719	0.03733	0.01604	157	-1.00	031.41	0
031	Phosphorus (%)	0065	0.78300	0.00000	0.81719	0.03733	0.01604	157	-0.92	031.01	0
031	Phosphorus (%)	0039	0.78315	0.00290	0.81719	0.03733	0.01604	157	-0.91	031.01	0
031	Phosphorus (%)	0695	0.78500	0.01000	0.81719	0.03733	0.01604	157	-0.86	031.41	0
031	Phosphorus (%)	2004	0.78550	0.01300	0.81719	0.03733	0.01604	157	-0.85	031.44	0
031	Phosphorus (%)	0045	0.78600	0.02800	0.81719	0.03733	0.01604	157	-0.84	031.42	0
031	Phosphorus (%)	0035	0.78645	0.01510	0.81719	0.03733	0.01604	157	-0.82	031.42	0
031	Phosphorus (%)	2023	0.78650	0.01700	0.81719	0.03733	0.01604	157	-0.82	031.44	0
031	Phosphorus (%)	0504	0.78800	0.00200	0.81719	0.03733	0.01604	157	-0.78	031.42	0
031	Phosphorus (%)	0619	0.78850	0.00500	0.81719	0.03733	0.01604	157	-0.77	031.41	0
031	Phosphorus (%)	0674	0.79000	0.02000	0.81719	0.03733	0.01604	157	-0.73	031.01	0
031	Phosphorus (%)	0171	0.79000	0.02000	0.81719	0.03733	0.01604	157	-0.73	031.41	0
031	Phosphorus (%)	0676	0.79000	0.00000	0.81719	0.03733	0.01604	157	-0.73	031.99	0
031	Phosphorus (%)	0553	0.79350	0.00100	0.81719	0.03733	0.01604	157	-0.63	031.41	0
031	Phosphorus (%)	0358	0.79500	0.05000	0.81719	0.03733	0.01604	157	-0.59	031.41	0
031	Phosphorus (%)	0175	0.79500	0.01000	0.81719	0.03733	0.01604	157	-0.59	031.01	0
031	Phosphorus (%)	0683	0.79500	0.01000	0.81719	0.03733	0.01604	157	-0.59	031.01	0
031	Phosphorus (%)	0629	0.79500	0.01000	0.81719	0.03733	0.01604	157	-0.59	031.43	0
031	Phosphorus (%)	0208	0.79550	0.02100	0.81719	0.03733	0.01604	157	-0.58	031.01	0
031	Phosphorus (%)	0033	0.79650	0.03700	0.81719	0.03733	0.01604	157	-0.55	031.43	0
031	Phosphorus (%)	0407	0.79700	0.00600	0.81719	0.03733	0.01604	157	-0.54	031.41	0
031	Phosphorus (%)	0233	0.80000	0.00000	0.81719	0.03733	0.01604	157	-0.46	031.01	0
031	Phosphorus (%)	0629	0.80000	0.02000	0.81719	0.03733	0.01604	157	-0.46	031.01	0
031	Phosphorus (%)	0019	0.80000	0.04000	0.81719	0.03733	0.01604	157	-0.46	031.41	0
031	Phosphorus (%)	2012	0.80000	0.00000	0.81719	0.03733	0.01604	157	-0.46	031.41	0
031	Phosphorus (%)	0297	0.80000	0.02000	0.81719	0.03733	0.01604	157	-0.46	031.43	0
031	Phosphorus (%)	0590	0.80000	0.00000	0.81719	0.03733	0.01604	157	-0.46	031.99	0
031	Phosphorus (%)	0918	0.80400	0.02200	0.81719	0.03733	0.01604	157	-0.35	031.43	0
031	Phosphorus (%)	0572	0.80470	0.05420	0.81719	0.03733	0.01604	157	-0.33	031.53	0
031	Phosphorus (%)	2011	0.80500	0.03000	0.81719	0.03733	0.01604	157	-0.33	031.01	0
031	Phosphorus (%)	0354	0.80500	0.01000	0.81719	0.03733	0.01604	157	-0.33	031.01	0
031	Phosphorus (%)	0897	0.80500	0.01000	0.81719	0.03733	0.01604	157	-0.33	031.01	0
031	Phosphorus (%)	2058	0.80500	0.01000	0.81719	0.03733	0.01604	157	-0.33	031.01	0



Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
031	Phosphorus (%)	0018	0.80650	0.04500	0.81719	0.03733	0.01604	157	-0.29	031.01	0
031	Phosphorus (%)	0891	0.80920	0.00140	0.81719	0.03733	0.01604	157	-0.21	031.01	0
031	Phosphorus (%)	0152	0.81000	0.00000	0.81719	0.03733	0.01604	157	-0.19	031.01	0
031	Phosphorus (%)	0878	0.81000	0.02000	0.81719	0.03733	0.01604	157	-0.19	031.01	0
031	Phosphorus (%)	2022	0.81000	0.00000	0.81719	0.03733	0.01604	157	-0.19	031.01	0
031	Phosphorus (%)	0049	0.81000	0.02000	0.81719	0.03733	0.01604	157	-0.19	031.41	0
031	Phosphorus (%)	0405	0.81000	0.02000	0.81719	0.03733	0.01604	157	-0.19	031.41	0
031	Phosphorus (%)	0263	0.81100	0.00060	0.81719	0.03733	0.01604	157	-0.17	031.01	0
031	Phosphorus (%)	0620	0.81175	0.00090	0.81719	0.03733	0.01604	157	-0.15	031.01	0
031	Phosphorus (%)	0011	0.81485	0.00190	0.81719	0.03733	0.01604	157	-0.06	031.02	0
031	Phosphorus (%)	2068	0.81500	0.01000	0.81719	0.03733	0.01604	157	-0.06	031.01	0
031	Phosphorus (%)	0536	0.81500	0.01000	0.81719	0.03733	0.01604	157	-0.06	031.06	0
031	Phosphorus (%)	0083	0.81500	0.01000	0.81719	0.03733	0.01604	157	-0.06	031.41	0
031	Phosphorus (%)	0190	0.81500	0.01000	0.81719	0.03733	0.01604	157	-0.06	031.42	0
031	Phosphorus (%)	0651	0.81550	0.00300	0.81719	0.03733	0.01604	157	-0.05	031.01	0
031	Phosphorus (%)	0001	0.81750	0.00300	0.81719	0.03733	0.01604	157	0.01	031.03	0
031	Phosphorus (%)	0849	0.81950	0.00100	0.81719	0.03733	0.01604	157	0.06	031.01	0
031	Phosphorus (%)	0142	0.82000	0.00000	0.81719	0.03733	0.01604	157	0.08	031.01	0
031	Phosphorus (%)	2042	0.82000	0.00000	0.81719	0.03733	0.01604	157	0.08	031.01	0
031	Phosphorus (%)	0043	0.82000	0.00000	0.81719	0.03733	0.01604	157	0.08	031.02	0
031	Phosphorus (%)	2053	0.82000	0.00000	0.81719	0.03733	0.01604	157	0.08	031.02	0
031	Phosphorus (%)	0552	0.82000	0.00000	0.81719	0.03733	0.01604	157	0.08	031.99	0
031	Phosphorus (%)	0884	0.82000	0.02000	0.81719	0.03733	0.01604	157	0.08	031.01	0
031	Phosphorus (%)	0100	0.82000	0.02000	0.81719	0.03733	0.01604	157	0.08	031.41	0
031	Phosphorus (%)	0353	0.82000	0.02000	0.81719	0.03733	0.01604	157	0.08	031.43	0
031	Phosphorus (%)	0723	0.82120	0.00000	0.81719	0.03733	0.01604	157	0.11	031.01	0
031	Phosphorus (%)	0138	0.82400	0.01200	0.81719	0.03733	0.01604	157	0.18	031.06	0
031	Phosphorus (%)	0937	0.82500	0.01000	0.81719	0.03733	0.01604	157	0.21	031.01	0
031	Phosphorus (%)	0074	0.82500	0.01000	0.81719	0.03733	0.01604	157	0.21	031.41	0
031	Phosphorus (%)	0148	0.82500	0.01000	0.81719	0.03733	0.01604	157	0.21	031.41	0
031	Phosphorus (%)	0037	0.82500	0.01000	0.81719	0.03733	0.01604	157	0.21	031.43	0
031	Phosphorus (%)	0008	0.82550	0.01100	0.81719	0.03733	0.01604	157	0.22	031.43	0
031	Phosphorus (%)	0038	0.82650	0.02700	0.81719	0.03733	0.01604	157	0.25	031.01	0
031	Phosphorus (%)	2062	0.82650	0.00500	0.81719	0.03733	0.01604	157	0.25	031.01	0
031	Phosphorus (%)	0300	0.82650	0.01300	0.81719	0.03733	0.01604	157	0.25	031.41	0
031	Phosphorus (%)	0187	0.82760	0.00040	0.81719	0.03733	0.01604	157	0.28	031.42	0
031	Phosphorus (%)	2065	0.82835	0.00330	0.81719	0.03733	0.01604	157	0.30	031.01	0
031	Phosphorus (%)	0726	0.82920	0.00140	0.81719	0.03733	0.01604	157	0.32	031.42	0
031	Phosphorus (%)	0229	0.83000	0.02000	0.81719	0.03733	0.01604	157	0.34	031.41	0
031	Phosphorus (%)	0202	0.83000	0.00000	0.81719	0.03733	0.01604	157	0.34	031.42	0
031	Phosphorus (%)	0357	0.83000	0.00000	0.81719	0.03733	0.01604	157	0.34	031.42	0
031	Phosphorus (%)	0653	0.83200	0.00400	0.81719	0.03733	0.01604	157	0.40	031.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
031	Phosphorus (%)	0205	0.83250	0.01700	0.81719	0.03733	0.01604	157	0.41	031.42	0
031	Phosphorus (%)	0968	0.83350	0.00300	0.81719	0.03733	0.01604	157	0.44	031.43	0
031	Phosphorus (%)	0036	0.83480	0.00620	0.81719	0.03733	0.01604	157	0.47	031.43	0
031	Phosphorus (%)	0893	0.83500	0.01000	0.81719	0.03733	0.01604	157	0.48	031.01	0
031	Phosphorus (%)	2063	0.83500	0.01000	0.81719	0.03733	0.01604	157	0.48	031.01	0
031	Phosphorus (%)	0004	0.83500	0.01000	0.81719	0.03733	0.01604	157	0.48	031.41	0
031	Phosphorus (%)	0164	0.83500	0.03000	0.81719	0.03733	0.01604	157	0.48	031.41	0
031	Phosphorus (%)	0425	0.83500	0.01000	0.81719	0.03733	0.01604	157	0.48	031.43	0
031	Phosphorus (%)	0036	0.83675	0.00170	0.81719	0.03733	0.01604	157	0.52	031.03	0
031	Phosphorus (%)	0208	0.83750	0.06100	0.81719	0.03733	0.01604	157	0.54	031.41	0
031	Phosphorus (%)	0626	0.83850	0.01500	0.81719	0.03733	0.01604	157	0.57	031.01	0
031	Phosphorus (%)	0011	0.83995	0.00810	0.81719	0.03733	0.01604	157	0.61	031.41	0
031	Phosphorus (%)	0178	0.84000	0.02000	0.81719	0.03733	0.01604	157	0.61	031.01	0
031	Phosphorus (%)	0505	0.84000	0.00000	0.81719	0.03733	0.01604	157	0.61	031.02	0
031	Phosphorus (%)	0720	0.84000	0.04000	0.81719	0.03733	0.01604	157	0.61	031.41	0
031	Phosphorus (%)	0675	0.84000	0.00000	0.81719	0.03733	0.01604	157	0.61	031.43	0
031	Phosphorus (%)	0939	0.84000	0.06000	0.81719	0.03733	0.01604	157	0.61	031.01	0
031	Phosphorus (%)	0014	0.84300	0.03000	0.81719	0.03733	0.01604	157	0.69	031.42	0
031	Phosphorus (%)	0038	0.84450	0.05100	0.81719	0.03733	0.01604	157	0.73	031.43	0
031	Phosphorus (%)	0687	0.84500	0.01000	0.81719	0.03733	0.01604	157	0.75	031.01	0
031	Phosphorus (%)	0098	0.84500	0.01000	0.81719	0.03733	0.01604	157	0.75	031.41	0
031	Phosphorus (%)	0881	0.84800	0.00800	0.81719	0.03733	0.01604	157	0.83	031.01	0
031	Phosphorus (%)	0868	0.84900	0.00200	0.81719	0.03733	0.01604	157	0.85	031.01	0
031	Phosphorus (%)	0051	0.84960	0.01020	0.81719	0.03733	0.01604	157	0.87	031.41	0
031	Phosphorus (%)	0609	0.85000	0.02000	0.81719	0.03733	0.01604	157	0.88	031.01	0
031	Phosphorus (%)	0948	0.85000	0.02000	0.81719	0.03733	0.01604	157	0.88	031.01	0
031	Phosphorus (%)	0895	0.85500	0.01000	0.81719	0.03733	0.01604	157	1.01	031.00	0
031	Phosphorus (%)	0194	0.85500	0.01000	0.81719	0.03733	0.01604	157	1.01	031.01	0
031	Phosphorus (%)	0903	0.85500	0.01000	0.81719	0.03733	0.01604	157	1.01	031.01	0
031	Phosphorus (%)	0028	0.85500	0.01000	0.81719	0.03733	0.01604	157	1.01	031.43	0
031	Phosphorus (%)	0560	0.85655	0.00850	0.81719	0.03733	0.01604	157	1.05	031.42	0
031	Phosphorus (%)	2006	0.85850	0.00500	0.81719	0.03733	0.01604	157	1.11	031.01	0
031	Phosphorus (%)	0689	0.86000	0.02000	0.81719	0.03733	0.01604	157	1.15	031.01	0
031	Phosphorus (%)	0278	0.86000	0.04000	0.81719	0.03733	0.01604	157	1.15	031.42	0
031	Phosphorus (%)	0623	0.86075	0.04470	0.81719	0.03733	0.01604	157	1.17	031.00	0
031	Phosphorus (%)	0964	0.86120	0.06960	0.81719	0.03733	0.01604	157	1.18	031.43	0
031	Phosphorus (%)	0345	0.86350	0.02900	0.81719	0.03733	0.01604	157	1.24	031.43	0
031	Phosphorus (%)	0921	0.86500	0.00000	0.81719	0.03733	0.01604	157	1.28	031.01	0
031	Phosphorus (%)	0413	0.86500	0.03000	0.81719	0.03733	0.01604	157	1.28	031.41	0
031	Phosphorus (%)	0199	0.86500	0.03000	0.81719	0.03733	0.01604	157	1.28	031.53	0
031	Phosphorus (%)	0598	0.86685	0.03510	0.81719	0.03733	0.01604	157	1.33	031.41	0
031	Phosphorus (%)	0610	0.87000	0.02000	0.81719	0.03733	0.01604	157	1.41	031.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
031	Phosphorus (%)	0889	0.87350	0.00500	0.81719	0.03733	0.01604	157	1.51	031.99	0
031	Phosphorus (%)	0964	0.87435	0.00130	0.81719	0.03733	0.01604	157	1.53	031.41	0
031	Phosphorus (%)	0003	0.87500	0.05000	0.81719	0.03733	0.01604	157	1.55	031.41	0
031	Phosphorus (%)	0682	0.87500	0.01000	0.81719	0.03733	0.01604	157	1.55	031.41	0
031	Phosphorus (%)	0226	0.88000	0.00000	0.81719	0.03733	0.01604	157	1.68	031.41	0
031	Phosphorus (%)	2067	0.88450	0.00100	0.81719	0.03733	0.01604	157	1.80	031.01	0
031	Phosphorus (%)	0555	0.88725	0.00850	0.81719	0.03733	0.01604	157	1.88	031.42	0
031	Phosphorus (%)	0563	0.89165	0.00590	0.81719	0.03733	0.01604	157	1.99	031.01	0
031	Phosphorus (%)	0106	0.91000	0.00000	0.81719	0.03733	0.01604	157	2.49	031.42	0
031	Phosphorus (%)	0520	0.91500	0.05000	0.81719	0.03733	0.01604	157	2.62	031.41	0
031	Phosphorus (%)	0098	0.92500	0.07000	0.81719	0.03733	0.01604	157	2.89	031.43	0
031	Phosphorus (%)	0066	1.0700	0.02000	0.81719	0.03733	0.01604	157	6.77	031.01	0
031	Phosphorus (%)	0265	0.84500	0.11000	0.81719	0.03733	0.01604	157	0.75	031.42	1
031	Phosphorus (%)	0337	0.96500	0.17000	0.81719	0.03733	0.01604	157	3.96	031.01	1
031	Phosphorus (%)	0047	0.00000	0.00000	0.81719	0.03733	0.01604	157	-21.89	031.03	2
031	Phosphorus (%)	0168	0.18950	0.22100	0.81719	0.03733	0.01604	157	-16.82	031.43	2
032	Potassium (%)	0921	0.68750	0.04500	1.0154	0.07807	0.02160	87	-4.20	032.31	0
032	Potassium (%)	0889	0.80750	0.01900	1.0154	0.07807	0.02160	87	-2.66	032.99	0
032	Potassium (%)	0049	0.83500	0.07000	1.0154	0.07807	0.02160	87	-2.31	032.41	0
032	Potassium (%)	0720	0.86500	0.01000	1.0154	0.07807	0.02160	87	-1.93	032.41	0
032	Potassium (%)	0139	0.87450	0.02500	1.0154	0.07807	0.02160	87	-1.80	032.02	0
032	Potassium (%)	0609	0.87500	0.01000	1.0154	0.07807	0.02160	87	-1.80	032.32	0
032	Potassium (%)	0536	0.88000	0.06000	1.0154	0.07807	0.02160	87	-1.73	032.31	0
032	Potassium (%)	0405	0.88500	0.03000	1.0154	0.07807	0.02160	87	-1.67	032.41	0
032	Potassium (%)	0003	0.89000	0.00000	1.0154	0.07807	0.02160	87	-1.61	032.41	0
032	Potassium (%)	0619	0.90050	0.02100	1.0154	0.07807	0.02160	87	-1.47	032.41	0
032	Potassium (%)	0881	0.90700	0.00600	1.0154	0.07807	0.02160	87	-1.39	032.02	0
032	Potassium (%)	0354	0.91000	0.00000	1.0154	0.07807	0.02160	87	-1.35	032.31	0
032	Potassium (%)	0510	0.91000	0.02000	1.0154	0.07807	0.02160	87	-1.35	032.43	0
032	Potassium (%)	2018	0.91500	0.03000	1.0154	0.07807	0.02160	87	-1.29	032.02	0
032	Potassium (%)	0154	0.93175	0.04290	1.0154	0.07807	0.02160	87	-1.07	032.52	0
032	Potassium (%)	0208	0.94400	0.06600	1.0154	0.07807	0.02160	87	-0.91	032.41	0
032	Potassium (%)	0968	0.94400	0.00400	1.0154	0.07807	0.02160	87	-0.91	032.43	0
032	Potassium (%)	0629	0.96000	0.00000	1.0154	0.07807	0.02160	87	-0.71	032.43	0
032	Potassium (%)	0874	0.96250	0.02080	1.0154	0.07807	0.02160	87	-0.68	032.31	0
032	Potassium (%)	0629	0.96500	0.01000	1.0154	0.07807	0.02160	87	-0.65	032.31	0
032	Potassium (%)	0353	0.96500	0.03000	1.0154	0.07807	0.02160	87	-0.65	032.43	0
032	Potassium (%)	0171	0.96500	0.09000	1.0154	0.07807	0.02160	87	-0.65	032.41	0
032	Potassium (%)	0142	0.96600	0.01200	1.0154	0.07807	0.02160	87	-0.63	032.31	0
032	Potassium (%)	0175	0.97500	0.05000	1.0154	0.07807	0.02160	87	-0.52	032.31	0
032	Potassium (%)	0144	0.97500	0.01000	1.0154	0.07807	0.02160	87	-0.52	032.41	0
032	Potassium (%)	2012	0.97500	0.05000	1.0154	0.07807	0.02160	87	-0.52	032.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
032	Potassium (%)	0948	0.98000	0.02000	1.0154	0.07807	0.02160	87	-0.45	032.31	0
032	Potassium (%)	0043	0.99000	0.00000	1.0154	0.07807	0.02160	87	-0.33	032.43	0
032	Potassium (%)	0242	0.99000	0.02000	1.0154	0.07807	0.02160	87	-0.33	032.99	0
032	Potassium (%)	0108	0.99500	0.03000	1.0154	0.07807	0.02160	87	-0.26	032.02	0
032	Potassium (%)	0358	0.99500	0.03000	1.0154	0.07807	0.02160	87	-0.26	032.41	0
032	Potassium (%)	0692	0.99500	0.01000	1.0154	0.07807	0.02160	87	-0.26	032.42	0
032	Potassium (%)	2023	0.99550	0.02900	1.0154	0.07807	0.02160	87	-0.25	032.44	0
032	Potassium (%)	0638	1.0000	0.00000	1.0154	0.07807	0.02160	87	-0.20	032.31	0
032	Potassium (%)	0683	1.0000	0.04000	1.0154	0.07807	0.02160	87	-0.20	032.31	0
032	Potassium (%)	0563	1.0008	0.01840	1.0154	0.07807	0.02160	87	-0.19	032.31	0
032	Potassium (%)	0065	1.0073	0.00650	1.0154	0.07807	0.02160	87	-0.10	032.31	0
032	Potassium (%)	0011	1.0109	0.02910	1.0154	0.07807	0.02160	87	-0.06	032.41	0
032	Potassium (%)	0598	1.0112	0.08160	1.0154	0.07807	0.02160	87	-0.05	032.41	0
032	Potassium (%)	0148	1.0150	0.01000	1.0154	0.07807	0.02160	87	0.00	032.41	0
032	Potassium (%)	0693	1.0150	0.03000	1.0154	0.07807	0.02160	87	0.00	032.42	0
032	Potassium (%)	0689	1.0200	0.00000	1.0154	0.07807	0.02160	87	0.06	032.31	0
032	Potassium (%)	0229	1.0200	0.00000	1.0154	0.07807	0.02160	87	0.06	032.41	0
032	Potassium (%)	0407	1.0200	0.02000	1.0154	0.07807	0.02160	87	0.06	032.41	0
032	Potassium (%)	0045	1.0205	0.05900	1.0154	0.07807	0.02160	87	0.07	032.42	0
032	Potassium (%)	0918	1.0235	0.01300	1.0154	0.07807	0.02160	87	0.10	032.43	0
032	Potassium (%)	0695	1.0300	0.02000	1.0154	0.07807	0.02160	87	0.19	032.41	0
032	Potassium (%)	0042	1.0300	0.02000	1.0154	0.07807	0.02160	87	0.19	032.42	0
032	Potassium (%)	0357	1.0300	0.00000	1.0154	0.07807	0.02160	87	0.19	032.42	0
032	Potassium (%)	0038	1.0350	0.03000	1.0154	0.07807	0.02160	87	0.25	032.31	0
032	Potassium (%)	0868	1.0350	0.01000	1.0154	0.07807	0.02160	87	0.25	032.31	0
032	Potassium (%)	0208	1.0400	0.00000	1.0154	0.07807	0.02160	87	0.32	032.31	0
032	Potassium (%)	0083	1.0400	0.02000	1.0154	0.07807	0.02160	87	0.32	032.41	0
032	Potassium (%)	0100	1.0400	0.02000	1.0154	0.07807	0.02160	87	0.32	032.41	0
032	Potassium (%)	0964	1.0449	0.00550	1.0154	0.07807	0.02160	87	0.38	032.41	0
032	Potassium (%)	0345	1.0450	0.01000	1.0154	0.07807	0.02160	87	0.38	032.43	0
032	Potassium (%)	2004	1.0450	0.01000	1.0154	0.07807	0.02160	87	0.38	032.44	0
032	Potassium (%)	0035	1.0470	0.00000	1.0154	0.07807	0.02160	87	0.40	032.42	0
032	Potassium (%)	0038	1.0500	0.04000	1.0154	0.07807	0.02160	87	0.44	032.43	0
032	Potassium (%)	2006	1.0515	0.00700	1.0154	0.07807	0.02160	87	0.46	032.08	0
032	Potassium (%)	0205	1.0550	0.03000	1.0154	0.07807	0.02160	87	0.51	032.42	0
032	Potassium (%)	0037	1.0600	0.00000	1.0154	0.07807	0.02160	87	0.57	032.43	0
032	Potassium (%)	0008	1.0610	0.03600	1.0154	0.07807	0.02160	87	0.58	032.43	0
032	Potassium (%)	0650	1.0650	0.01000	1.0154	0.07807	0.02160	87	0.64	032.31	0
032	Potassium (%)	0675	1.0650	0.01000	1.0154	0.07807	0.02160	87	0.64	032.43	0
032	Potassium (%)	0187	1.0700	0.00000	1.0154	0.07807	0.02160	87	0.70	032.42	0
032	Potassium (%)	0164	1.0750	0.01000	1.0154	0.07807	0.02160	87	0.76	032.41	0
032	Potassium (%)	0202	1.0750	0.01000	1.0154	0.07807	0.02160	87	0.76	032.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
032	Potassium (%)	0199	1.0750	0.01000	1.0154	0.07807	0.02160	87	0.76	032.53	0
032	Potassium (%)	0413	1.0800	0.02000	1.0154	0.07807	0.02160	87	0.83	032.41	0
032	Potassium (%)	0297	1.0800	0.00000	1.0154	0.07807	0.02160	87	0.83	032.43	0
032	Potassium (%)	0504	1.0810	0.00800	1.0154	0.07807	0.02160	87	0.84	032.42	0
032	Potassium (%)	0033	1.0900	0.08000	1.0154	0.07807	0.02160	87	0.96	032.43	0
032	Potassium (%)	0504	1.0900	0.02000	1.0154	0.07807	0.02160	87	0.96	032.02	0
032	Potassium (%)	0098	1.0900	0.02000	1.0154	0.07807	0.02160	87	0.96	032.41	0
032	Potassium (%)	0682	1.0900	0.00000	1.0154	0.07807	0.02160	87	0.96	032.41	0
032	Potassium (%)	0294	1.0950	0.01000	1.0154	0.07807	0.02160	87	1.02	032.42	0
032	Potassium (%)	0226	1.1000	0.02000	1.0154	0.07807	0.02160	87	1.08	032.41	0
032	Potassium (%)	0878	1.1000	0.00000	1.0154	0.07807	0.02160	87	1.08	032.41	0
032	Potassium (%)	0560	1.1033	0.00950	1.0154	0.07807	0.02160	87	1.13	032.42	0
032	Potassium (%)	0425	1.1100	0.02000	1.0154	0.07807	0.02160	87	1.21	032.43	0
032	Potassium (%)	0572	1.1285	0.00700	1.0154	0.07807	0.02160	87	1.45	032.53	0
032	Potassium (%)	0035	1.1395	0.03300	1.0154	0.07807	0.02160	87	1.59	032.32	0
032	Potassium (%)	0106	1.1400	0.02000	1.0154	0.07807	0.02160	87	1.60	032.42	0
032	Potassium (%)	0098	1.1400	0.08000	1.0154	0.07807	0.02160	87	1.60	032.43	0
032	Potassium (%)	0555	1.1777	0.01530	1.0154	0.07807	0.02160	87	2.08	032.42	0
032	Potassium (%)	0520	1.1850	0.03000	1.0154	0.07807	0.02160	87	2.17	032.41	0
032	Potassium (%)	0278	1.1000	0.12000	1.0154	0.07807	0.02160	87	1.08	032.42	1
032	Potassium (%)	0964	1.1024	0.11760	1.0154	0.07807	0.02160	87	1.11	032.43	1
032	Potassium (%)	0265	1.1250	0.13000	1.0154	0.07807	0.02160	87	1.40	032.42	1
032	Potassium (%)	0051	0.23170	0.00640	1.0154	0.07807	0.02160	87	-10.04	032.41	2
033	Salt as chloride (%)	0910	0.00000	0.00000	0.59619	0.07057	0.01081	93	-8.45	033.00	0
033	Salt as chloride (%)	0891	0.18150	0.00160	0.59619	0.07057	0.01081	93	-5.88	033.00	0
033	Salt as chloride (%)	0623	0.18570	0.02280	0.59619	0.07057	0.01081	93	-5.82	033.99	0
033	Salt as chloride (%)	2012	0.23000	0.00000	0.59619	0.07057	0.01081	93	-5.19	033.00	0
033	Salt as chloride (%)	2022	0.30600	0.00000	0.59619	0.07057	0.01081	93	-4.11	033.00	0
033	Salt as chloride (%)	2068	0.33000	0.00000	0.59619	0.07057	0.01081	93	-3.77	033.00	0
033	Salt as chloride (%)	2058	0.37000	0.02000	0.59619	0.07057	0.01081	93	-3.20	033.00	0
033	Salt as chloride (%)	0685	0.39500	0.03000	0.59619	0.07057	0.01081	93	-2.85	033.00	0
033	Salt as chloride (%)	2011	0.43850	0.01100	0.59619	0.07057	0.01081	93	-2.23	033.01	0
033	Salt as chloride (%)	0505	0.48000	0.02000	0.59619	0.07057	0.01081	93	-1.65	033.03	0
033	Salt as chloride (%)	0536	0.48000	0.02000	0.59619	0.07057	0.01081	93	-1.65	033.99	0
033	Salt as chloride (%)	2006	0.48700	0.01000	0.59619	0.07057	0.01081	93	-1.55	033.00	0
033	Salt as chloride (%)	0265	0.49000	0.00000	0.59619	0.07057	0.01081	93	-1.50	033.03	0
033	Salt as chloride (%)	0681	0.49000	0.04000	0.59619	0.07057	0.01081	93	-1.50	033.99	0
033	Salt as chloride (%)	0895	0.49500	0.01000	0.59619	0.07057	0.01081	93	-1.43	033.00	0
033	Salt as chloride (%)	0726	0.50000	0.00000	0.59619	0.07057	0.01081	93	-1.36	033.03	0
033	Salt as chloride (%)	0903	0.50500	0.01000	0.59619	0.07057	0.01081	93	-1.29	033.00	0
033	Salt as chloride (%)	2063	0.51000	0.02000	0.59619	0.07057	0.01081	93	-1.22	033.00	0
033	Salt as chloride (%)	0552	0.51000	0.00000	0.59619	0.07057	0.01081	93	-1.22	033.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
033	Salt as chloride (%)	0881	0.51350	0.00100	0.59619	0.07057	0.01081	93	-1.17	033.00	0
033	Salt as chloride (%)	2065	0.51500	0.00340	0.59619	0.07057	0.01081	93	-1.15	033.00	0
033	Salt as chloride (%)	0298	0.52000	0.02000	0.59619	0.07057	0.01081	93	-1.08	033.00	0
033	Salt as chloride (%)	0848	0.52000	0.02000	0.59619	0.07057	0.01081	93	-1.08	033.03	0
033	Salt as chloride (%)	0358	0.52500	0.03000	0.59619	0.07057	0.01081	93	-1.01	033.99	0
033	Salt as chloride (%)	0893	0.53500	0.01000	0.59619	0.07057	0.01081	93	-0.87	033.00	0
033	Salt as chloride (%)	0017	0.54000	0.02000	0.59619	0.07057	0.01081	93	-0.80	033.99	0
033	Salt as chloride (%)	0539	0.54500	0.03000	0.59619	0.07057	0.01081	93	-0.73	033.00	0
033	Salt as chloride (%)	0868	0.55255	0.01150	0.59619	0.07057	0.01081	93	-0.62	033.00	0
033	Salt as chloride (%)	0849	0.55500	0.03000	0.59619	0.07057	0.01081	93	-0.58	033.00	0
033	Salt as chloride (%)	0278	0.55500	0.01000	0.59619	0.07057	0.01081	93	-0.58	033.01	0
033	Salt as chloride (%)	0723	0.56000	0.00000	0.59619	0.07057	0.01081	93	-0.51	033.00	0
033	Salt as chloride (%)	0937	0.56000	0.00000	0.59619	0.07057	0.01081	93	-0.51	033.05	0
033	Salt as chloride (%)	0885	0.56500	0.01000	0.59619	0.07057	0.01081	93	-0.44	033.00	0
033	Salt as chloride (%)	2042	0.57000	0.00000	0.59619	0.07057	0.01081	93	-0.37	033.00	0
033	Salt as chloride (%)	0004	0.57000	0.00000	0.59619	0.07057	0.01081	93	-0.37	033.05	0
033	Salt as chloride (%)	0878	0.58000	0.02000	0.59619	0.07057	0.01081	93	-0.23	033.00	0
033	Salt as chloride (%)	0190	0.58000	0.04000	0.59619	0.07057	0.01081	93	-0.23	033.03	0
033	Salt as chloride (%)	2006	0.58900	0.00800	0.59619	0.07057	0.01081	93	-0.10	033.01	0
033	Salt as chloride (%)	2006	0.58900	0.00800	0.59619	0.07057	0.01081	93	-0.10	033.05	0
033	Salt as chloride (%)	0884	0.59500	0.01000	0.59619	0.07057	0.01081	93	-0.02	033.00	0
033	Salt as chloride (%)	0309	0.59850	0.00340	0.59619	0.07057	0.01081	93	0.03	033.00	0
033	Salt as chloride (%)	0629	0.60000	0.00000	0.59619	0.07057	0.01081	93	0.05	033.01	0
033	Salt as chloride (%)	2062	0.60450	0.00900	0.59619	0.07057	0.01081	93	0.12	033.00	0
033	Salt as chloride (%)	0083	0.60500	0.01000	0.59619	0.07057	0.01081	93	0.12	033.01	0
033	Salt as chloride (%)	0511	0.61000	0.00000	0.59619	0.07057	0.01081	93	0.20	033.00	0
033	Salt as chloride (%)	0638	0.61000	0.00000	0.59619	0.07057	0.01081	93	0.20	033.01	0
033	Salt as chloride (%)	0939	0.61000	0.00000	0.59619	0.07057	0.01081	93	0.20	033.01	0
033	Salt as chloride (%)	0693	0.61500	0.05000	0.59619	0.07057	0.01081	93	0.27	033.00	0
033	Salt as chloride (%)	0619	0.61650	0.01100	0.59619	0.07057	0.01081	93	0.29	033.99	0
033	Salt as chloride (%)	0178	0.62350	0.00300	0.59619	0.07057	0.01081	93	0.39	033.01	0
033	Salt as chloride (%)	0683	0.62500	0.03000	0.59619	0.07057	0.01081	93	0.41	033.00	0
033	Salt as chloride (%)	0001	0.62500	0.01000	0.59619	0.07057	0.01081	93	0.41	033.01	0
033	Salt as chloride (%)	0229	0.62500	0.01000	0.59619	0.07057	0.01081	93	0.41	033.01	0
033	Salt as chloride (%)	0337	0.62500	0.01000	0.59619	0.07057	0.01081	93	0.41	033.01	0
033	Salt as chloride (%)	0682	0.62500	0.01000	0.59619	0.07057	0.01081	93	0.41	033.01	0
033	Salt as chloride (%)	0011	0.62580	0.01440	0.59619	0.07057	0.01081	93	0.42	033.01	0
033	Salt as chloride (%)	0964	0.62890	0.01240	0.59619	0.07057	0.01081	93	0.46	033.01	0
033	Salt as chloride (%)	0948	0.63000	0.00000	0.59619	0.07057	0.01081	93	0.48	033.01	0
033	Salt as chloride (%)	2023	0.63000	0.00000	0.59619	0.07057	0.01081	93	0.48	033.01	0
033	Salt as chloride (%)	0171	0.63000	0.04000	0.59619	0.07057	0.01081	93	0.48	033.05	0
033	Salt as chloride (%)	0148	0.63050	0.00300	0.59619	0.07057	0.01081	93	0.49	033.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
033	Salt as chloride (%)	0098	0.63500	0.01000	0.59619	0.07057	0.01081	93	0.55	033.01	0
033	Salt as chloride (%)	0226	0.63500	0.01000	0.59619	0.07057	0.01081	93	0.55	033.01	0
033	Salt as chloride (%)	0683	0.63500	0.01000	0.59619	0.07057	0.01081	93	0.55	033.01	0
033	Salt as chloride (%)	0626	0.63600	0.00400	0.59619	0.07057	0.01081	93	0.56	033.99	0
033	Salt as chloride (%)	0407	0.63900	0.00400	0.59619	0.07057	0.01081	93	0.61	033.01	0
033	Salt as chloride (%)	0897	0.64000	0.00000	0.59619	0.07057	0.01081	93	0.62	033.00	0
033	Salt as chloride (%)	0100	0.64000	0.00000	0.59619	0.07057	0.01081	93	0.62	033.01	0
033	Salt as chloride (%)	0164	0.64000	0.00000	0.59619	0.07057	0.01081	93	0.62	033.01	0
033	Salt as chloride (%)	0354	0.64000	0.00000	0.59619	0.07057	0.01081	93	0.62	033.01	0
033	Salt as chloride (%)	0413	0.64000	0.02000	0.59619	0.07057	0.01081	93	0.62	033.01	0
033	Salt as chloride (%)	0510	0.64000	0.00000	0.59619	0.07057	0.01081	93	0.62	033.01	0
033	Salt as chloride (%)	0689	0.64000	0.00000	0.59619	0.07057	0.01081	93	0.62	033.05	0
033	Salt as chloride (%)	0205	0.64300	0.00800	0.59619	0.07057	0.01081	93	0.66	033.01	0
033	Salt as chloride (%)	0353	0.64500	0.01000	0.59619	0.07057	0.01081	93	0.69	033.00	0
033	Salt as chloride (%)	0590	0.64500	0.03000	0.59619	0.07057	0.01081	93	0.69	033.01	0
033	Salt as chloride (%)	0175	0.65000	0.02000	0.59619	0.07057	0.01081	93	0.76	033.01	0
033	Salt as chloride (%)	0425	0.65000	0.00000	0.59619	0.07057	0.01081	93	0.76	033.01	0
033	Salt as chloride (%)	0039	0.65420	0.00500	0.59619	0.07057	0.01081	93	0.82	033.01	0
033	Salt as chloride (%)	0106	0.65500	0.00000	0.59619	0.07057	0.01081	93	0.83	033.01	0
033	Salt as chloride (%)	0242	0.66000	0.00000	0.59619	0.07057	0.01081	93	0.90	033.01	0
033	Salt as chloride (%)	0874	0.66060	0.00160	0.59619	0.07057	0.01081	93	0.91	033.01	0
033	Salt as chloride (%)	0194	0.66500	0.01000	0.59619	0.07057	0.01081	93	0.98	033.01	0
033	Salt as chloride (%)	0650	0.67000	0.02000	0.59619	0.07057	0.01081	93	1.05	033.01	0
033	Salt as chloride (%)	0208	0.68000	0.04200	0.59619	0.07057	0.01081	93	1.19	033.99	0
033	Salt as chloride (%)	0169	0.68000	0.00000	0.59619	0.07057	0.01081	93	1.19	033.00	0
033	Salt as chloride (%)	0674	0.70000	0.00000	0.59619	0.07057	0.01081	93	1.47	033.01	0
033	Salt as chloride (%)	0675	0.70500	0.01000	0.59619	0.07057	0.01081	93	1.54	033.00	0
033	Salt as chloride (%)	0045	0.70950	0.00700	0.59619	0.07057	0.01081	93	1.61	033.00	0
033	Salt as chloride (%)	0861	0.71000	0.00000	0.59619	0.07057	0.01081	93	1.61	033.01	0
033	Salt as chloride (%)	2053	0.72500	0.01000	0.59619	0.07057	0.01081	93	1.83	033.01	0
033	Salt as chloride (%)	0019	0.73000	0.02000	0.59619	0.07057	0.01081	93	1.90	033.01	0
033	Salt as chloride (%)	2043	0.75500	0.03000	0.59619	0.07057	0.01081	93	2.25	033.05	0
033	Salt as chloride (%)	0144	0.48500	0.07000	0.59619	0.07057	0.01081	93	-1.58	033.03	1
033	Salt as chloride (%)	0003	0.65000	0.10000	0.59619	0.07057	0.01081	93	0.76	033.01	1
034	Selenium (mg / kg (ppm))	0425	0.00000	0.00000	0.44988	0.09155	0.07159	27	-4.91	034.43	0
034	Selenium (mg / kg (ppm))	0190	0.22500	0.01000	0.44988	0.09155	0.07159	27	-2.46	034.99	0
034	Selenium (mg / kg (ppm))	0693	0.29500	0.05000	0.44988	0.09155	0.07159	27	-1.69	034.42	0
034	Selenium (mg / kg (ppm))	2055	0.35000	0.10000	0.44988	0.09155	0.07159	27	-1.09	034.53	0
034	Selenium (mg / kg (ppm))	0003	0.36000	0.48000	0.44988	0.09155	0.07159	27	-0.98	034.43	0
034	Selenium (mg / kg (ppm))	0169	0.37000	0.00000	0.44988	0.09155	0.07159	27	-0.87	034.04	0
034	Selenium (mg / kg (ppm))	0939	0.37500	0.03000	0.44988	0.09155	0.07159	27	-0.82	034.04	0
034	Selenium (mg / kg (ppm))	0171	0.38150	0.00100	0.44988	0.09155	0.07159	27	-0.75	034.04	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
034	Selenium (mg / kg (ppm))	0199	0.40000	0.00000	0.44988	0.09155	0.07159	27	-0.54	034.53	0
034	Selenium (mg / kg (ppm))	0563	0.41150	0.02300	0.44988	0.09155	0.07159	27	-0.42	034.04	0
034	Selenium (mg / kg (ppm))	0638	0.42000	0.00000	0.44988	0.09155	0.07159	27	-0.33	034.01	0
034	Selenium (mg / kg (ppm))	2053	0.45000	0.02000	0.44988	0.09155	0.07159	27	0.00	034.04	0
034	Selenium (mg / kg (ppm))	2058	0.45500	0.01000	0.44988	0.09155	0.07159	27	0.06	034.53	0
034	Selenium (mg / kg (ppm))	0619	0.45950	0.00700	0.44988	0.09155	0.07159	27	0.11	034.41	0
034	Selenium (mg / kg (ppm))	0038	0.47000	0.02600	0.44988	0.09155	0.07159	27	0.22	034.01	0
034	Selenium (mg / kg (ppm))	0033	0.47000	0.06000	0.44988	0.09155	0.07159	27	0.22	034.43	0
034	Selenium (mg / kg (ppm))	0098	0.48000	0.02000	0.44988	0.09155	0.07159	27	0.33	034.01	0
034	Selenium (mg / kg (ppm))	0164	0.48000	0.02000	0.44988	0.09155	0.07159	27	0.33	034.04	0
034	Selenium (mg / kg (ppm))	0208	0.48100	0.05600	0.44988	0.09155	0.07159	27	0.34	034.04	0
034	Selenium (mg / kg (ppm))	0553	0.50500	0.03400	0.44988	0.09155	0.07159	27	0.60	034.53	0
034	Selenium (mg / kg (ppm))	0555	0.51000	0.12000	0.44988	0.09155	0.07159	27	0.66	034.99	0
034	Selenium (mg / kg (ppm))	0610	0.52000	0.00000	0.44988	0.09155	0.07159	27	0.77	034.04	0
034	Selenium (mg / kg (ppm))	0047	0.54000	0.02000	0.44988	0.09155	0.07159	27	0.98	034.52	0
034	Selenium (mg / kg (ppm))	2023	0.55000	0.10000	0.44988	0.09155	0.07159	27	1.09	034.53	0
034	Selenium (mg / kg (ppm))	0098	0.55900	0.08400	0.44988	0.09155	0.07159	27	1.19	034.52	0
034	Selenium (mg / kg (ppm))	0918	0.62000	0.17200	0.44988	0.09155	0.07159	27	1.86	034.53	0
034	Selenium (mg / kg (ppm))	2067	2.6450	0.49000	0.44988	0.09155	0.07159	27	23.98	034.31	0
034	Selenium (mg / kg (ppm))	0168	12.870	4.7000	0.44988	0.09155	0.07159	27	135.66	034.53	2
035	Sodium (%)	2055	0.16900	0.00400	0.22240	0.01385	0.00740	99	-3.86	035.99	0
035	Sodium (%)	0089	0.19000	0.00000	0.22240	0.01385	0.00740	99	-2.34	035.41	0
035	Sodium (%)	0175	0.19500	0.01000	0.22240	0.01385	0.00740	99	-1.98	035.31	0
035	Sodium (%)	0609	0.19500	0.01000	0.22240	0.01385	0.00740	99	-1.98	035.32	0
035	Sodium (%)	0510	0.19600	0.00800	0.22240	0.01385	0.00740	99	-1.91	035.43	0
035	Sodium (%)	0139	0.19850	0.00300	0.22240	0.01385	0.00740	99	-1.73	035.05	0
035	Sodium (%)	2018	0.20000	0.02000	0.22240	0.01385	0.00740	99	-1.62	035.05	0
035	Sodium (%)	0298	0.20000	0.02000	0.22240	0.01385	0.00740	99	-1.62	035.41	0
035	Sodium (%)	0353	0.20000	0.02000	0.22240	0.01385	0.00740	99	-1.62	035.43	0
035	Sodium (%)	0154	0.20015	0.01490	0.22240	0.01385	0.00740	99	-1.61	035.52	0
035	Sodium (%)	0049	0.20500	0.01000	0.22240	0.01385	0.00740	99	-1.26	035.41	0
035	Sodium (%)	0171	0.20500	0.01000	0.22240	0.01385	0.00740	99	-1.26	035.41	0
035	Sodium (%)	0358	0.20500	0.01000	0.22240	0.01385	0.00740	99	-1.26	035.41	0
035	Sodium (%)	0035	0.20525	0.00230	0.22240	0.01385	0.00740	99	-1.24	035.42	0
035	Sodium (%)	0619	0.20650	0.00900	0.22240	0.01385	0.00740	99	-1.15	035.41	0
035	Sodium (%)	0144	0.21000	0.00000	0.22240	0.01385	0.00740	99	-0.90	035.41	0
035	Sodium (%)	0297	0.21000	0.00000	0.22240	0.01385	0.00740	99	-0.90	035.43	0
035	Sodium (%)	0208	0.21050	0.01300	0.22240	0.01385	0.00740	99	-0.86	035.41	0
035	Sodium (%)	0011	0.21060	0.00860	0.22240	0.01385	0.00740	99	-0.85	035.41	0
035	Sodium (%)	2004	0.21100	0.01600	0.22240	0.01385	0.00740	99	-0.82	035.41	0
035	Sodium (%)	0918	0.21200	0.00400	0.22240	0.01385	0.00740	99	-0.75	035.43	0
035	Sodium (%)	0598	0.21315	0.00830	0.22240	0.01385	0.00740	99	-0.67	035.41	0



Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
035	Sodium (%)	0921	0.21400	0.00400	0.22240	0.01385	0.00740	99	-0.61	035.31	0
035	Sodium (%)	0045	0.21400	0.00600	0.22240	0.01385	0.00740	99	-0.61	035.42	0
035	Sodium (%)	0683	0.21500	0.01000	0.22240	0.01385	0.00740	99	-0.53	035.31	0
035	Sodium (%)	0242	0.21500	0.01000	0.22240	0.01385	0.00740	99	-0.53	035.99	0
035	Sodium (%)	0951	0.21500	0.03000	0.22240	0.01385	0.00740	99	-0.53	035.05	0
035	Sodium (%)	0300	0.21550	0.00100	0.22240	0.01385	0.00740	99	-0.50	035.41	0
035	Sodium (%)	0563	0.21605	0.00130	0.22240	0.01385	0.00740	99	-0.46	035.31	0
035	Sodium (%)	0868	0.21700	0.00400	0.22240	0.01385	0.00740	99	-0.39	035.01	0
035	Sodium (%)	0208	0.21750	0.01900	0.22240	0.01385	0.00740	99	-0.35	035.31	0
035	Sodium (%)	0407	0.21750	0.00500	0.22240	0.01385	0.00740	99	-0.35	035.41	0
035	Sodium (%)	0138	0.21800	0.00400	0.22240	0.01385	0.00740	99	-0.32	035.01	0
035	Sodium (%)	0038	0.21800	0.00200	0.22240	0.01385	0.00740	99	-0.32	035.43	0
035	Sodium (%)	2006	0.21850	0.00100	0.22240	0.01385	0.00740	99	-0.28	035.01	0
035	Sodium (%)	0638	0.21850	0.00100	0.22240	0.01385	0.00740	99	-0.28	035.31	0
035	Sodium (%)	0038	0.21950	0.00100	0.22240	0.01385	0.00740	99	-0.21	035.31	0
035	Sodium (%)	0345	0.21950	0.00500	0.22240	0.01385	0.00740	99	-0.21	035.43	0
035	Sodium (%)	0964	0.21955	0.00090	0.22240	0.01385	0.00740	99	-0.21	035.41	0
035	Sodium (%)	0233	0.22000	0.00000	0.22240	0.01385	0.00740	99	-0.17	035.31	0
035	Sodium (%)	0629	0.22000	0.00000	0.22240	0.01385	0.00740	99	-0.17	035.31	0
035	Sodium (%)	0650	0.22000	0.02000	0.22240	0.01385	0.00740	99	-0.17	035.31	0
035	Sodium (%)	0689	0.22000	0.00000	0.22240	0.01385	0.00740	99	-0.17	035.31	0
035	Sodium (%)	0948	0.22000	0.02000	0.22240	0.01385	0.00740	99	-0.17	035.31	0
035	Sodium (%)	0520	0.22000	0.02000	0.22240	0.01385	0.00740	99	-0.17	035.41	0
035	Sodium (%)	0682	0.22000	0.00000	0.22240	0.01385	0.00740	99	-0.17	035.41	0
035	Sodium (%)	2012	0.22000	0.00000	0.22240	0.01385	0.00740	99	-0.17	035.41	0
035	Sodium (%)	0692	0.22000	0.00000	0.22240	0.01385	0.00740	99	-0.17	035.42	0
035	Sodium (%)	0693	0.22000	0.00000	0.22240	0.01385	0.00740	99	-0.17	035.42	0
035	Sodium (%)	0881	0.22050	0.00100	0.22240	0.01385	0.00740	99	-0.14	035.05	0
035	Sodium (%)	0008	0.22050	0.01100	0.22240	0.01385	0.00740	99	-0.14	035.43	0
035	Sodium (%)	0042	0.22150	0.01100	0.22240	0.01385	0.00740	99	-0.07	035.42	0
035	Sodium (%)	0148	0.22200	0.00400	0.22240	0.01385	0.00740	99	-0.03	035.41	0
035	Sodium (%)	2023	0.22300	0.00000	0.22240	0.01385	0.00740	99	0.04	035.41	0
035	Sodium (%)	0205	0.22350	0.00900	0.22240	0.01385	0.00740	99	0.08	035.42	0
035	Sodium (%)	0553	0.22450	0.00100	0.22240	0.01385	0.00740	99	0.15	035.41	0
035	Sodium (%)	0354	0.22500	0.01000	0.22240	0.01385	0.00740	99	0.19	035.31	0
035	Sodium (%)	0100	0.22500	0.01000	0.22240	0.01385	0.00740	99	0.19	035.41	0
035	Sodium (%)	0164	0.22500	0.01000	0.22240	0.01385	0.00740	99	0.19	035.41	0
035	Sodium (%)	0294	0.22500	0.01000	0.22240	0.01385	0.00740	99	0.19	035.42	0
035	Sodium (%)	0425	0.22500	0.01000	0.22240	0.01385	0.00740	99	0.19	035.43	0
035	Sodium (%)	2058	0.22500	0.01000	0.22240	0.01385	0.00740	99	0.19	035.43	0
035	Sodium (%)	0505	0.22550	0.00700	0.22240	0.01385	0.00740	99	0.22	035.31	0
035	Sodium (%)	0142	0.22600	0.00200	0.22240	0.01385	0.00740	99	0.26	035.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
035	Sodium (%)	0629	0.22600	0.00000	0.22240	0.01385	0.00740	99	0.26	035.43	0
035	Sodium (%)	0874	0.22670	0.00080	0.22240	0.01385	0.00740	99	0.31	035.31	0
035	Sodium (%)	0035	0.22930	0.00160	0.22240	0.01385	0.00740	99	0.50	035.32	0
035	Sodium (%)	0108	0.23000	0.02000	0.22240	0.01385	0.00740	99	0.55	035.05	0
035	Sodium (%)	0265	0.23000	0.02000	0.22240	0.01385	0.00740	99	0.55	035.42	0
035	Sodium (%)	0536	0.23000	0.00000	0.22240	0.01385	0.00740	99	0.55	035.31	0
035	Sodium (%)	0083	0.23000	0.00000	0.22240	0.01385	0.00740	99	0.55	035.41	0
035	Sodium (%)	0229	0.23000	0.00000	0.22240	0.01385	0.00740	99	0.55	035.41	0
035	Sodium (%)	0878	0.23000	0.00000	0.22240	0.01385	0.00740	99	0.55	035.41	0
035	Sodium (%)	0202	0.23000	0.00000	0.22240	0.01385	0.00740	99	0.55	035.42	0
035	Sodium (%)	0033	0.23250	0.02100	0.22240	0.01385	0.00740	99	0.73	035.43	0
035	Sodium (%)	0065	0.23260	0.00280	0.22240	0.01385	0.00740	99	0.74	035.31	0
035	Sodium (%)	0405	0.23300	0.00400	0.22240	0.01385	0.00740	99	0.77	035.41	0
035	Sodium (%)	0572	0.23315	0.01130	0.22240	0.01385	0.00740	99	0.78	035.53	0
035	Sodium (%)	0152	0.23350	0.00700	0.22240	0.01385	0.00740	99	0.80	035.05	0
035	Sodium (%)	0337	0.23500	0.01000	0.22240	0.01385	0.00740	99	0.91	035.05	0
035	Sodium (%)	0504	0.23500	0.01000	0.22240	0.01385	0.00740	99	0.91	035.05	0
035	Sodium (%)	0098	0.23500	0.01000	0.22240	0.01385	0.00740	99	0.91	035.41	0
035	Sodium (%)	0413	0.23500	0.01000	0.22240	0.01385	0.00740	99	0.91	035.41	0
035	Sodium (%)	0504	0.23700	0.02200	0.22240	0.01385	0.00740	99	1.05	035.42	0
035	Sodium (%)	0560	0.23745	0.00410	0.22240	0.01385	0.00740	99	1.09	035.42	0
035	Sodium (%)	0098	0.24000	0.02000	0.22240	0.01385	0.00740	99	1.27	035.43	0
035	Sodium (%)	0964	0.24335	0.01170	0.22240	0.01385	0.00740	99	1.51	035.43	0
035	Sodium (%)	0226	0.24500	0.01000	0.22240	0.01385	0.00740	99	1.63	035.41	0
035	Sodium (%)	0695	0.25000	0.00000	0.22240	0.01385	0.00740	99	1.99	035.41	0
035	Sodium (%)	0199	0.25000	0.02000	0.22240	0.01385	0.00740	99	1.99	035.53	0
035	Sodium (%)	0187	0.25115	0.00030	0.22240	0.01385	0.00740	99	2.08	035.42	0
035	Sodium (%)	0868	0.25450	0.00900	0.22240	0.01385	0.00740	99	2.32	035.31	0
035	Sodium (%)	0720	0.25500	0.01000	0.22240	0.01385	0.00740	99	2.35	035.41	0
035	Sodium (%)	0675	0.25500	0.01000	0.22240	0.01385	0.00740	99	2.35	035.43	0
035	Sodium (%)	0278	0.26000	0.00000	0.22240	0.01385	0.00740	99	2.71	035.42	0
035	Sodium (%)	0555	0.26145	0.00330	0.22240	0.01385	0.00740	99	2.82	035.42	0
035	Sodium (%)	0263	0.27425	0.00150	0.22240	0.01385	0.00740	99	3.74	035.32	0
035	Sodium (%)	0106	0.30000	0.00000	0.22240	0.01385	0.00740	99	5.60	035.42	0
035	Sodium (%)	0889	0.32500	0.01000	0.22240	0.01385	0.00740	99	7.41	035.99	0
035	Sodium (%)	0043	0.65500	0.01000	0.22240	0.01385	0.00740	99	31.24	035.43	2
036	Sulfur (%)	0229	0.25500	0.01000	0.28099	0.02199	0.00809	30	-1.18	036.04	0
036	Sulfur (%)	0693	0.25500	0.01000	0.28099	0.02199	0.00809	30	-1.18	036.42	0
036	Sulfur (%)	0510	0.25500	0.01000	0.28099	0.02199	0.00809	30	-1.18	036.43	0
036	Sulfur (%)	0045	0.25850	0.01900	0.28099	0.02199	0.00809	30	-1.02	036.42	0
036	Sulfur (%)	0226	0.26000	0.00000	0.28099	0.02199	0.00809	30	-0.95	036.04	0
036	Sulfur (%)	0297	0.26000	0.00000	0.28099	0.02199	0.00809	30	-0.95	036.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
036	Sulfur (%)	0345	0.26250	0.00500	0.28099	0.02199	0.00809	30	-0.84	036.43	0
036	Sulfur (%)	0629	0.26500	0.00000	0.28099	0.02199	0.00809	30	-0.73	036.43	0
036	Sulfur (%)	0708	0.26850	0.00300	0.28099	0.02199	0.00809	30	-0.57	036.42	0
036	Sulfur (%)	0164	0.27000	0.02000	0.28099	0.02199	0.00809	30	-0.50	036.42	0
036	Sulfur (%)	0202	0.27000	0.00000	0.28099	0.02199	0.00809	30	-0.50	036.42	0
036	Sulfur (%)	0357	0.27000	0.00000	0.28099	0.02199	0.00809	30	-0.50	036.42	0
036	Sulfur (%)	0353	0.27000	0.02000	0.28099	0.02199	0.00809	30	-0.50	036.43	0
036	Sulfur (%)	0242	0.27000	0.00000	0.28099	0.02199	0.00809	30	-0.50	036.99	0
036	Sulfur (%)	0918	0.27050	0.00300	0.28099	0.02199	0.00809	30	-0.48	036.43	0
036	Sulfur (%)	0038	0.27800	0.01400	0.28099	0.02199	0.00809	30	-0.14	036.43	0
036	Sulfur (%)	0407	0.28250	0.02100	0.28099	0.02199	0.00809	30	0.07	036.42	0
036	Sulfur (%)	0294	0.28500	0.01000	0.28099	0.02199	0.00809	30	0.18	036.42	0
036	Sulfur (%)	0098	0.29000	0.00000	0.28099	0.02199	0.00809	30	0.41	036.04	0
036	Sulfur (%)	0278	0.29000	0.00000	0.28099	0.02199	0.00809	30	0.41	036.42	0
036	Sulfur (%)	0560	0.29320	0.00240	0.28099	0.02199	0.00809	30	0.56	036.42	0
036	Sulfur (%)	0171	0.29500	0.03000	0.28099	0.02199	0.00809	30	0.64	036.42	0
036	Sulfur (%)	0555	0.30100	0.00540	0.28099	0.02199	0.00809	30	0.91	036.42	0
036	Sulfur (%)	0187	0.30355	0.00090	0.28099	0.02199	0.00809	30	1.03	036.42	0
036	Sulfur (%)	0098	0.30500	0.03000	0.28099	0.02199	0.00809	30	1.09	036.43	0
036	Sulfur (%)	0169	0.31000	0.02000	0.28099	0.02199	0.00809	30	1.32	036.43	0
036	Sulfur (%)	0033	0.31400	0.00000	0.28099	0.02199	0.00809	30	1.50	036.43	0
036	Sulfur (%)	0106	0.32000	0.00000	0.28099	0.02199	0.00809	30	1.77	036.42	0
036	Sulfur (%)	0889	0.32550	0.00100	0.28099	0.02199	0.00809	30	2.02	036.99	0
036	Sulfur (%)	0042	0.32800	0.00800	0.28099	0.02199	0.00809	30	2.14	036.42	0
036	Sulfur (%)	0265	0.31000	0.06000	0.28099	0.02199	0.00809	30	1.32	036.42	1
037	Zinc (mg / kg (ppm))	0695	73.805	0.15000	95.809	7.3634	3.2571	98	-2.99	037.41	0
037	Zinc (mg / kg (ppm))	0199	77.000	2.0000	95.809	7.3634	3.2571	98	-2.55	037.53	0
037	Zinc (mg / kg (ppm))	0405	79.500	1.0000	95.809	7.3634	3.2571	98	-2.21	037.41	0
037	Zinc (mg / kg (ppm))	0164	83.500	3.0000	95.809	7.3634	3.2571	98	-1.67	037.41	0
037	Zinc (mg / kg (ppm))	0178	84.150	1.3000	95.809	7.3634	3.2571	98	-1.58	037.31	0
037	Zinc (mg / kg (ppm))	0619	84.250	1.5000	95.809	7.3634	3.2571	98	-1.57	037.41	0
037	Zinc (mg / kg (ppm))	0358	85.125	8.0900	95.809	7.3634	3.2571	98	-1.45	037.41	0
037	Zinc (mg / kg (ppm))	2022	85.200	0.20000	95.809	7.3634	3.2571	98	-1.44	037.31	0
037	Zinc (mg / kg (ppm))	2058	85.260	4.9200	95.809	7.3634	3.2571	98	-1.43	037.43	0
037	Zinc (mg / kg (ppm))	0175	86.000	12.000	95.809	7.3634	3.2571	98	-1.33	037.31	0
037	Zinc (mg / kg (ppm))	0049	86.095	4.9900	95.809	7.3634	3.2571	98	-1.32	037.41	0
037	Zinc (mg / kg (ppm))	0208	86.445	0.19000	95.809	7.3634	3.2571	98	-1.27	037.41	0
037	Zinc (mg / kg (ppm))	0242	86.500	3.0000	95.809	7.3634	3.2571	98	-1.26	037.99	0
037	Zinc (mg / kg (ppm))	2055	87.300	7.0000	95.809	7.3634	3.2571	98	-1.16	037.99	0
037	Zinc (mg / kg (ppm))	0874	87.500	5.0000	95.809	7.3634	3.2571	98	-1.13	037.31	0
037	Zinc (mg / kg (ppm))	0043	88.150	2.7000	95.809	7.3634	3.2571	98	-1.04	037.43	0
037	Zinc (mg / kg (ppm))	0098	88.200	0.00000	95.809	7.3634	3.2571	98	-1.03	037.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
037	Zinc (mg / kg (ppm))	0629	88.500	1.0000	95.809	7.3634	3.2571	98	-0.99	037.43	0
037	Zinc (mg / kg (ppm))	0918	88.770	1.6800	95.809	7.3634	3.2571	98	-0.96	037.43	0
037	Zinc (mg / kg (ppm))	0638	89.100	2.6000	95.809	7.3634	3.2571	98	-0.91	037.31	0
037	Zinc (mg / kg (ppm))	0004	89.500	5.0000	95.809	7.3634	3.2571	98	-0.86	037.41	0
037	Zinc (mg / kg (ppm))	0555	90.165	0.75370	95.809	7.3634	3.2571	98	-0.77	037.42	0
037	Zinc (mg / kg (ppm))	0510	90.500	3.0000	95.809	7.3634	3.2571	98	-0.72	037.43	0
037	Zinc (mg / kg (ppm))	0187	91.075	1.1900	95.809	7.3634	3.2571	98	-0.64	037.42	0
037	Zinc (mg / kg (ppm))	0171	91.450	5.7000	95.809	7.3634	3.2571	98	-0.59	037.41	0
037	Zinc (mg / kg (ppm))	0609	91.500	3.0000	95.809	7.3634	3.2571	98	-0.59	037.32	0
037	Zinc (mg / kg (ppm))	0511	91.500	3.0000	95.809	7.3634	3.2571	98	-0.59	037.41	0
037	Zinc (mg / kg (ppm))	0689	91.800	0.20000	95.809	7.3634	3.2571	98	-0.54	037.31	0
037	Zinc (mg / kg (ppm))	0610	92.000	2.0000	95.809	7.3634	3.2571	98	-0.52	037.41	0
037	Zinc (mg / kg (ppm))	2012	92.090	0.00000	95.809	7.3634	3.2571	98	-0.51	037.41	0
037	Zinc (mg / kg (ppm))	2023	92.200	1.0000	95.809	7.3634	3.2571	98	-0.49	037.43	0
037	Zinc (mg / kg (ppm))	0939	92.215	0.31000	95.809	7.3634	3.2571	98	-0.49	037.31	0
037	Zinc (mg / kg (ppm))	0723	92.500	0.00000	95.809	7.3634	3.2571	98	-0.45	037.31	0
037	Zinc (mg / kg (ppm))	0968	92.500	1.0000	95.809	7.3634	3.2571	98	-0.45	037.43	0
037	Zinc (mg / kg (ppm))	0653	92.575	1.5900	95.809	7.3634	3.2571	98	-0.44	037.31	0
037	Zinc (mg / kg (ppm))	0868	93.000	0.00000	95.809	7.3634	3.2571	98	-0.38	037.31	0
037	Zinc (mg / kg (ppm))	0948	93.025	1.1300	95.809	7.3634	3.2571	98	-0.38	037.33	0
037	Zinc (mg / kg (ppm))	0300	93.083	0.84800	95.809	7.3634	3.2571	98	-0.37	037.41	0
037	Zinc (mg / kg (ppm))	0168	93.500	13.000	95.809	7.3634	3.2571	98	-0.31	037.43	0
037	Zinc (mg / kg (ppm))	0598	94.100	4.3400	95.809	7.3634	3.2571	98	-0.23	037.41	0
037	Zinc (mg / kg (ppm))	2004	94.300	7.0000	95.809	7.3634	3.2571	98	-0.20	037.44	0
037	Zinc (mg / kg (ppm))	0148	94.450	0.10000	95.809	7.3634	3.2571	98	-0.18	037.41	0
037	Zinc (mg / kg (ppm))	0294	94.565	1.6300	95.809	7.3634	3.2571	98	-0.17	037.42	0
037	Zinc (mg / kg (ppm))	0354	94.650	0.08000	95.809	7.3634	3.2571	98	-0.16	037.31	0
037	Zinc (mg / kg (ppm))	0066	95.035	4.5500	95.809	7.3634	3.2571	98	-0.11	037.31	0
037	Zinc (mg / kg (ppm))	0345	95.300	1.6000	95.809	7.3634	3.2571	98	-0.07	037.43	0
037	Zinc (mg / kg (ppm))	0878	95.500	3.0000	95.809	7.3634	3.2571	98	-0.04	037.41	0
037	Zinc (mg / kg (ppm))	0889	95.500	1.0000	95.809	7.3634	3.2571	98	-0.04	037.99	0
037	Zinc (mg / kg (ppm))	0629	96.000	2.0000	95.809	7.3634	3.2571	98	0.03	037.31	0
037	Zinc (mg / kg (ppm))	0910	96.000	6.0000	95.809	7.3634	3.2571	98	0.03	037.41	0
037	Zinc (mg / kg (ppm))	0051	96.040	0.68000	95.809	7.3634	3.2571	98	0.03	037.41	0
037	Zinc (mg / kg (ppm))	0505	96.100	1.6000	95.809	7.3634	3.2571	98	0.04	037.31	0
037	Zinc (mg / kg (ppm))	0008	96.200	0.60000	95.809	7.3634	3.2571	98	0.05	037.43	0
037	Zinc (mg / kg (ppm))	0675	96.545	0.09000	95.809	7.3634	3.2571	98	0.10	037.43	0
037	Zinc (mg / kg (ppm))	0921	96.850	16.300	95.809	7.3634	3.2571	98	0.14	037.31	0
037	Zinc (mg / kg (ppm))	0553	96.900	3.6000	95.809	7.3634	3.2571	98	0.15	037.41	0
037	Zinc (mg / kg (ppm))	0563	97.000	2.0000	95.809	7.3634	3.2571	98	0.16	037.31	0
037	Zinc (mg / kg (ppm))	0045	97.300	13.400	95.809	7.3634	3.2571	98	0.20	037.42	0
037	Zinc (mg / kg (ppm))	0011	97.473	0.33500	95.809	7.3634	3.2571	98	0.23	037.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
037	Zinc (mg / kg (ppm))	0038	97.500	3.0000	95.809	7.3634	3.2571	98	0.23	037.32	0
037	Zinc (mg / kg (ppm))	0572	98.200	7.6000	95.809	7.3634	3.2571	98	0.32	037.53	0
037	Zinc (mg / kg (ppm))	0674	98.430	2.4600	95.809	7.3634	3.2571	98	0.36	037.31	0
037	Zinc (mg / kg (ppm))	0504	98.620	0.24000	95.809	7.3634	3.2571	98	0.38	037.33	0
037	Zinc (mg / kg (ppm))	0208	98.650	1.9000	95.809	7.3634	3.2571	98	0.39	037.31	0
037	Zinc (mg / kg (ppm))	2067	99.600	0.62000	95.809	7.3634	3.2571	98	0.51	037.31	0
037	Zinc (mg / kg (ppm))	0720	99.733	4.8980	95.809	7.3634	3.2571	98	0.53	037.41	0
037	Zinc (mg / kg (ppm))	0357	100.00	2.0000	95.809	7.3634	3.2571	98	0.57	037.42	0
037	Zinc (mg / kg (ppm))	0693	100.00	6.0000	95.809	7.3634	3.2571	98	0.57	037.42	0
037	Zinc (mg / kg (ppm))	0297	100.00	6.0000	95.809	7.3634	3.2571	98	0.57	037.43	0
037	Zinc (mg / kg (ppm))	0033	100.05	1.9000	95.809	7.3634	3.2571	98	0.58	037.43	0
037	Zinc (mg / kg (ppm))	0353	100.23	1.5400	95.809	7.3634	3.2571	98	0.60	037.43	0
037	Zinc (mg / kg (ppm))	0035	100.25	0.10000	95.809	7.3634	3.2571	98	0.60	037.32	0
037	Zinc (mg / kg (ppm))	0425	100.45	0.30000	95.809	7.3634	3.2571	98	0.63	037.43	0
037	Zinc (mg / kg (ppm))	0083	101.00	4.0000	95.809	7.3634	3.2571	98	0.71	037.41	0
037	Zinc (mg / kg (ppm))	0520	101.00	4.0000	95.809	7.3634	3.2571	98	0.71	037.41	0
037	Zinc (mg / kg (ppm))	0726	101.30	0.40000	95.809	7.3634	3.2571	98	0.75	037.42	0
037	Zinc (mg / kg (ppm))	0560	101.50	1.0000	95.809	7.3634	3.2571	98	0.77	037.42	0
037	Zinc (mg / kg (ppm))	0407	101.60	3.8400	95.809	7.3634	3.2571	98	0.79	037.41	0
037	Zinc (mg / kg (ppm))	0190	102.47	0.33000	95.809	7.3634	3.2571	98	0.90	037.42	0
037	Zinc (mg / kg (ppm))	2018	102.52	6.0400	95.809	7.3634	3.2571	98	0.91	037.31	0
037	Zinc (mg / kg (ppm))	0590	103.00	10.000	95.809	7.3634	3.2571	98	0.98	037.31	0
037	Zinc (mg / kg (ppm))	0003	103.50	3.0000	95.809	7.3634	3.2571	98	1.04	037.41	0
037	Zinc (mg / kg (ppm))	0226	103.50	5.0000	95.809	7.3634	3.2571	98	1.04	037.41	0
037	Zinc (mg / kg (ppm))	0098	103.85	4.7000	95.809	7.3634	3.2571	98	1.09	037.43	0
037	Zinc (mg / kg (ppm))	0100	104.00	4.0000	95.809	7.3634	3.2571	98	1.11	037.41	0
037	Zinc (mg / kg (ppm))	0035	104.15	4.3000	95.809	7.3634	3.2571	98	1.13	037.42	0
037	Zinc (mg / kg (ppm))	0964	104.39	5.1200	95.809	7.3634	3.2571	98	1.17	037.41	0
037	Zinc (mg / kg (ppm))	0682	104.83	4.2900	95.809	7.3634	3.2571	98	1.22	037.41	0
037	Zinc (mg / kg (ppm))	0229	105.00	2.0000	95.809	7.3634	3.2571	98	1.25	037.41	0
037	Zinc (mg / kg (ppm))	0504	105.50	1.4000	95.809	7.3634	3.2571	98	1.32	037.32	0
037	Zinc (mg / kg (ppm))	0169	105.50	1.0000	95.809	7.3634	3.2571	98	1.32	037.43	0
037	Zinc (mg / kg (ppm))	0413	106.00	2.0000	95.809	7.3634	3.2571	98	1.38	037.42	0
037	Zinc (mg / kg (ppm))	0964	106.60	14.830	95.809	7.3634	3.2571	98	1.46	037.43	0
037	Zinc (mg / kg (ppm))	0265	107.50	1.0000	95.809	7.3634	3.2571	98	1.59	037.42	0
037	Zinc (mg / kg (ppm))	0106	112.50	10.340	95.809	7.3634	3.2571	98	2.27	037.42	0
037	Zinc (mg / kg (ppm))	0042	114.00	8.0000	95.809	7.3634	3.2571	98	2.47	037.42	0
037	Zinc (mg / kg (ppm))	0202	114.01	2.0000	95.809	7.3634	3.2571	98	2.47	037.42	0
037	Zinc (mg / kg (ppm))	0536	138.75	3.1000	95.809	7.3634	3.2571	98	5.83	037.31	0
037	Zinc (mg / kg (ppm))	0692	101.00	22.000	95.809	7.3634	3.2571	98	0.71	037.42	1
037	Zinc (mg / kg (ppm))	0074	104.00	22.000	95.809	7.3634	3.2571	98	1.11	037.41	1
037	Zinc (mg / kg (ppm))	0951	107.15	23.300	95.809	7.3634	3.2571	98	1.54	037.31	1

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
037	Zinc (mg / kg (ppm))	0038	0.27800	0.01400	95.809	7.3634	3.2571	98	-12.97	037.43	2
038	Molybdenum (mg / kg (ppm))	0098	0.77100	0.18400	1.5697	0.58849	0.13472	18	-1.36	038.52	0
038	Molybdenum (mg / kg (ppm))	0297	0.89500	0.23000	1.5697	0.58849	0.13472	18	-1.15	038.43	0
038	Molybdenum (mg / kg (ppm))	0693	0.95500	0.39000	1.5697	0.58849	0.13472	18	-1.04	038.42	0
038	Molybdenum (mg / kg (ppm))	0560	0.96500	0.13000	1.5697	0.58849	0.13472	18	-1.03	038.42	0
038	Molybdenum (mg / kg (ppm))	0353	1.0500	0.04000	1.5697	0.58849	0.13472	18	-0.88	038.43	0
038	Molybdenum (mg / kg (ppm))	2055	1.2000	0.20000	1.5697	0.58849	0.13472	18	-0.63	038.99	0
038	Molybdenum (mg / kg (ppm))	0345	1.3750	0.05000	1.5697	0.58849	0.13472	18	-0.33	038.43	0
038	Molybdenum (mg / kg (ppm))	0918	1.4775	0.34500	1.5697	0.58849	0.13472	18	-0.16	038.53	0
038	Molybdenum (mg / kg (ppm))	0011	1.5325	0.07000	1.5697	0.58849	0.13472	18	-0.06	038.41	0
038	Molybdenum (mg / kg (ppm))	0171	1.5400	0.06000	1.5697	0.58849	0.13472	18	-0.05	038.41	0
038	Molybdenum (mg / kg (ppm))	0510	1.7500	0.10000	1.5697	0.58849	0.13472	18	0.31	038.43	0
038	Molybdenum (mg / kg (ppm))	0964	1.8198	0.09030	1.5697	0.58849	0.13472	18	0.42	038.41	0
038	Molybdenum (mg / kg (ppm))	0553	1.9950	0.03000	1.5697	0.58849	0.13472	18	0.72	038.53	0
038	Molybdenum (mg / kg (ppm))	0964	2.0511	0.09570	1.5697	0.58849	0.13472	18	0.82	038.43	0
038	Molybdenum (mg / kg (ppm))	2023	2.0750	0.13000	1.5697	0.58849	0.13472	18	0.86	038.53	0
038	Molybdenum (mg / kg (ppm))	0278	2.1000	0.14000	1.5697	0.58849	0.13472	18	0.90	038.42	0
038	Molybdenum (mg / kg (ppm))	0038	2.2500	0.10000	1.5697	0.58849	0.13472	18	1.16	038.43	0
038	Molybdenum (mg / kg (ppm))	0003	2.4600	0.04000	1.5697	0.58849	0.13472	18	1.51	038.43	0
038	Molybdenum (mg / kg (ppm))	0169	1.5250	0.59000	1.5697	0.58849	0.13472	18	-0.08	038.43	1
040	Barium (mg / kg (ppm))	0918	9.1175	0.22300	9.7225	0.67464	0.57433	3	-0.90	040.53	0
040	Barium (mg / kg (ppm))	2055	9.6000	0.60000	9.7225	0.67464	0.57433	3	-0.18	040.99	0
040	Barium (mg / kg (ppm))	0560	10.450	0.90000	9.7225	0.67464	0.57433	3	1.08	040.42	0
041	Vanadium (mg / kg (ppm))	0553	0.41600	0.03200			0.03200	1			
042	Chloride (%)	0297	0.35000	0.00000			0.00000	1			
057	Ethoxyquin (mg / kg (ppm))	0227	0.00000	0.00000	0.39167	0.61458	0.00333	3	-0.64	057.01	0
057	Ethoxyquin (mg / kg (ppm))	0626	0.07500	0.01000	0.39167	0.61458	0.00333	3	-0.52	057.00	0
057	Ethoxyquin (mg / kg (ppm))	0875	1.1000	0.00000	0.39167	0.61458	0.00333	3	1.15	057.01	0
065	Monensin (mg / kg (ppm))	2055	31.000	4.0000	41.422	2.3649	1.6852	20	-4.41	065.99	0
065	Monensin (mg / kg (ppm))	0001	35.045	0.19000	41.422	2.3649	1.6852	20	-2.70	065.03	0
065	Monensin (mg / kg (ppm))	0905	38.300	0.80000	41.422	2.3649	1.6852	20	-1.32	065.03	0
065	Monensin (mg / kg (ppm))	0619	38.700	0.00000	41.422	2.3649	1.6852	20	-1.15	065.99	0
065	Monensin (mg / kg (ppm))	0875	39.800	2.2000	41.422	2.3649	1.6852	20	-0.69	065.03	0
065	Monensin (mg / kg (ppm))	0668	40.050	0.90000	41.422	2.3649	1.6852	20	-0.58	065.03	0
065	Monensin (mg / kg (ppm))	0047	40.400	0.80000	41.422	2.3649	1.6852	20	-0.43	065.04	0
065	Monensin (mg / kg (ppm))	0171	41.300	0.00000	41.422	2.3649	1.6852	20	-0.05	065.99	0
065	Monensin (mg / kg (ppm))	0036	41.550	0.10000	41.422	2.3649	1.6852	20	0.05	065.04	0
065	Monensin (mg / kg (ppm))	2053	41.600	1.2000	41.422	2.3649	1.6852	20	0.08	065.03	0
065	Monensin (mg / kg (ppm))	0033	41.600	0.80000	41.422	2.3649	1.6852	20	0.08	065.99	0
065	Monensin (mg / kg (ppm))	0043	41.750	0.50000	41.422	2.3649	1.6852	20	0.14	065.00	0
065	Monensin (mg / kg (ppm))	0032	41.750	1.9000	41.422	2.3649	1.6852	20	0.14	065.04	0
065	Monensin (mg / kg (ppm))	0668	42.000	0.00000	41.422	2.3649	1.6852	20	0.24	065.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
065	Monensin (mg / kg (ppm))	0218	42.274	1.3130	41.422	2.3649	1.6852	20	0.36	065.04	0
065	Monensin (mg / kg (ppm))	0006	43.000	2.0000	41.422	2.3649	1.6852	20	0.67	065.04	0
065	Monensin (mg / kg (ppm))	0148	43.940	3.6000	41.422	2.3649	1.6852	20	1.06	065.03	0
065	Monensin (mg / kg (ppm))	2004	44.000	0.00000	41.422	2.3649	1.6852	20	1.09	065.04	0
065	Monensin (mg / kg (ppm))	0010	48.900	6.6000	41.422	2.3649	1.6852	20	3.16	065.99	0
065	Monensin (mg / kg (ppm))	0013	51.400	6.8000	41.422	2.3649	1.6852	20	4.22	065.03	0
102	Niacin (mg / kg (ppm))	0227	108.50	9.0000			9.0000	1			
103	Pantothenic Acid (mg / kg (ppm))	0227	9.7750	0.85000			0.85000	1			
104	Riboflavin (mg / kg (ppm))	0227	2.7850	0.17000	3.1175	0.47023	0.13500	2	-0.71	104.00	0
104	Riboflavin (mg / kg (ppm))	2023	3.4500	0.10000	3.1175	0.47023	0.13500	2	0.71	104.03	0
105	Thiamine (mg / kg (ppm))	2023	7.4150	0.05000			0.05000	1			
106	Vitamin A (KU / kg)	0910	0.00000	0.00000	2.8231	1.5156	0.29300	21	-1.86	106.02	0
106	Vitamin A (KU / kg)	2023	1.1950	0.39000	2.8231	1.5156	0.29300	21	-1.07	106.02	0
106	Vitamin A (KU / kg)	0921	1.3900	0.00000	2.8231	1.5156	0.29300	21	-0.95	106.02	0
106	Vitamin A (KU / kg)	0555	1.4500	0.50000	2.8231	1.5156	0.29300	21	-0.91	106.02	0
106	Vitamin A (KU / kg)	0169	1.5500	0.10000	2.8231	1.5156	0.29300	21	-0.84	106.02	0
106	Vitamin A (KU / kg)	0638	1.6100	0.00000	2.8231	1.5156	0.29300	21	-0.80	106.02	0
106	Vitamin A (KU / kg)	0905	1.7850	0.31200	2.8231	1.5156	0.29300	21	-0.68	106.02	0
106	Vitamin A (KU / kg)	0171	2.1750	0.05000	2.8231	1.5156	0.29300	21	-0.43	106.00	0
106	Vitamin A (KU / kg)	0676	2.4000	0.60000	2.8231	1.5156	0.29300	21	-0.28	106.02	0
106	Vitamin A (KU / kg)	0861	2.6450	0.07000	2.8231	1.5156	0.29300	21	-0.12	106.02	0
106	Vitamin A (KU / kg)	0563	2.8502	0.24910	2.8231	1.5156	0.29300	21	0.02	106.02	0
106	Vitamin A (KU / kg)	2058	3.0000	0.12000	2.8231	1.5156	0.29300	21	0.12	106.02	0
106	Vitamin A (KU / kg)	2053	3.2500	0.72000	2.8231	1.5156	0.29300	21	0.28	106.02	0
106	Vitamin A (KU / kg)	0941	3.3935	0.51700	2.8231	1.5156	0.29300	21	0.38	106.02	0
106	Vitamin A (KU / kg)	0208	3.5270	0.13200	2.8231	1.5156	0.29300	21	0.46	106.02	0
106	Vitamin A (KU / kg)	0098	3.5505	0.03900	2.8231	1.5156	0.29300	21	0.48	106.01	0
106	Vitamin A (KU / kg)	0610	3.7000	0.80000	2.8231	1.5156	0.29300	21	0.58	106.02	0
106	Vitamin A (KU / kg)	0227	3.9750	0.33000	2.8231	1.5156	0.29300	21	0.76	106.02	0
106	Vitamin A (KU / kg)	0675	5.3500	0.10000	2.8231	1.5156	0.29300	21	1.67	106.02	0
106	Vitamin A (KU / kg)	2067	6.3310	0.02400	2.8231	1.5156	0.29300	21	2.31	106.02	0
106	Vitamin A (KU / kg)	0689	9.1500	1.1000	2.8231	1.5156	0.29300	21	4.17	106.02	0
106	Vitamin A (KU / kg)	0004	3.9025	1.4990	2.8231	1.5156	0.29300	21	0.71	106.02	1
107	Vitamin B12 (µg / kg (ppb))	0227	12.250	0.10000			0.10000	1			
108	Vitamin D3 (KU / kg)	2023	0.24500	0.03000	0.97000	0.76830	0.08000	4	-0.94	108.01	0
108	Vitamin D3 (KU / kg)	2058	0.75000	0.08000	0.97000	0.76830	0.08000	4	-0.29	108.02	0
108	Vitamin D3 (KU / kg)	0675	0.83000	0.04000	0.97000	0.76830	0.08000	4	-0.18	108.02	0
108	Vitamin D3 (KU / kg)	0227	2.0550	0.17000	0.97000	0.76830	0.08000	4	1.41	108.02	0
109	Vitamin E (mg / kg (ppm))	0098	0.00000	0.00000	20.610	11.653	0.71806	18	-1.77	109.02	0
109	Vitamin E (mg / kg (ppm))	0910	0.00000	0.00000	20.610	11.653	0.71806	18	-1.77	109.02	0
109	Vitamin E (mg / kg (ppm))	0555	11.185	0.13000	20.610	11.653	0.71806	18	-0.81	109.02	0
109	Vitamin E (mg / kg (ppm))	0955	11.550	0.90000	20.610	11.653	0.71806	18	-0.78	109.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
109	Vitamin E (mg / kg (ppm))	2058	14.745	0.19000	20.610	11.653	0.71806	18	-0.50	109.02	0
109	Vitamin E (mg / kg (ppm))	0921	16.200	0.00000	20.610	11.653	0.71806	18	-0.38	109.02	0
109	Vitamin E (mg / kg (ppm))	0676	16.700	0.40000	20.610	11.653	0.71806	18	-0.34	109.02	0
109	Vitamin E (mg / kg (ppm))	0905	16.775	1.3900	20.610	11.653	0.71806	18	-0.33	109.02	0
109	Vitamin E (mg / kg (ppm))	0208	17.310	0.00000	20.610	11.653	0.71806	18	-0.28	109.02	0
109	Vitamin E (mg / kg (ppm))	0227	19.300	1.2000	20.610	11.653	0.71806	18	-0.11	109.02	0
109	Vitamin E (mg / kg (ppm))	0861	20.000	2.0000	20.610	11.653	0.71806	18	-0.05	109.02	0
109	Vitamin E (mg / kg (ppm))	0638	25.450	0.00000	20.610	11.653	0.71806	18	0.42	109.02	0
109	Vitamin E (mg / kg (ppm))	2023	28.350	0.30000	20.610	11.653	0.71806	18	0.66	109.02	0
109	Vitamin E (mg / kg (ppm))	0563	30.475	0.36500	20.610	11.653	0.71806	18	0.85	109.02	0
109	Vitamin E (mg / kg (ppm))	0941	37.385	1.0900	20.610	11.653	0.71806	18	1.44	109.02	0
109	Vitamin E (mg / kg (ppm))	0675	39.245	0.59000	20.610	11.653	0.71806	18	1.60	109.02	0
109	Vitamin E (mg / kg (ppm))	2067	50.500	1.0000	20.610	11.653	0.71806	18	2.57	109.02	0
109	Vitamin E (mg / kg (ppm))	0199	53.485	3.3700	20.610	11.653	0.71806	18	2.82	109.02	0
109	Vitamin E (mg / kg (ppm))	0610	26.750	5.5000	20.610	11.653	0.71806	18	0.53	109.02	1
110	Xanthophyll (mg / kg (ppm))	0955	8.3800	0.48000			0.48000	1			
113	Folic Acid (mg / kg (ppm))	0227	0.91450	0.04900			0.04900	1			
114	Biotin (mg / kg (ppm))	0227	0.26500	0.00400			0.00400	1			
120	Alanine (%)	2022	0.79500	0.03000	0.86114	0.02102	0.01446	20	-3.15	120.00	0
120	Alanine (%)	0297	0.80450	0.00300	0.86114	0.02102	0.01446	20	-2.69	120.01	0
120	Alanine (%)	0504	0.84000	0.02000	0.86114	0.02102	0.01446	20	-1.01	120.00	0
120	Alanine (%)	2023	0.84000	0.00000	0.86114	0.02102	0.01446	20	-1.01	120.02	0
120	Alanine (%)	0619	0.84000	0.02740	0.86114	0.02102	0.01446	20	-1.01	120.00	0
120	Alanine (%)	0644	0.84650	0.00100	0.86114	0.02102	0.01446	20	-0.70	120.00	0
120	Alanine (%)	0675	0.85500	0.01000	0.86114	0.02102	0.01446	20	-0.29	120.00	0
120	Alanine (%)	0968	0.85800	0.00200	0.86114	0.02102	0.01446	20	-0.15	120.00	0
120	Alanine (%)	2004	0.85950	0.00100	0.86114	0.02102	0.01446	20	-0.08	120.00	0
120	Alanine (%)	2059	0.85950	0.00900	0.86114	0.02102	0.01446	20	-0.08	120.00	0
120	Alanine (%)	0571	0.86150	0.00900	0.86114	0.02102	0.01446	20	0.02	120.00	0
120	Alanine (%)	0872	0.86540	0.01240	0.86114	0.02102	0.01446	20	0.20	120.00	0
120	Alanine (%)	0684	0.86950	0.02100	0.86114	0.02102	0.01446	20	0.40	120.00	0
120	Alanine (%)	0227	0.87000	0.02000	0.86114	0.02102	0.01446	20	0.42	120.00	0
120	Alanine (%)	0652	0.87000	0.02000	0.86114	0.02102	0.01446	20	0.42	120.00	0
120	Alanine (%)	0948	0.87200	0.01000	0.86114	0.02102	0.01446	20	0.52	120.01	0
120	Alanine (%)	0676	0.88500	0.05000	0.86114	0.02102	0.01446	20	1.14	120.05	0
120	Alanine (%)	0098	0.89300	0.00600	0.86114	0.02102	0.01446	20	1.52	120.02	0
120	Alanine (%)	0918	0.89450	0.02900	0.86114	0.02102	0.01446	20	1.59	120.00	0
120	Alanine (%)	0868	0.91385	0.00830	0.86114	0.02102	0.01446	20	2.51	120.00	0
120	Alanine (%)	0008	0.81050	0.11100	0.86114	0.02102	0.01446	20	-2.41	120.05	1
121	Arginine (%)	0948	0.75100	0.03600	0.86631	0.03425	0.01509	21	-3.37	121.01	0
121	Arginine (%)	0504	0.80500	0.01000	0.86631	0.03425	0.01509	21	-1.79	121.00	0
121	Arginine (%)	0227	0.81000	0.00000	0.86631	0.03425	0.01509	21	-1.64	121.00	0



Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
121	Arginine (%)	0619	0.82400	0.01400	0.86631	0.03425	0.01509	21	-1.24	121.00	0
121	Arginine (%)	0684	0.84450	0.00100	0.86631	0.03425	0.01509	21	-0.64	121.00	0
121	Arginine (%)	0675	0.84500	0.01000	0.86631	0.03425	0.01509	21	-0.62	121.00	0
121	Arginine (%)	0968	0.85400	0.00800	0.86631	0.03425	0.01509	21	-0.36	121.00	0
121	Arginine (%)	0098	0.85950	0.00500	0.86631	0.03425	0.01509	21	-0.20	121.02	0
121	Arginine (%)	0872	0.86215	0.00070	0.86631	0.03425	0.01509	21	-0.12	121.00	0
121	Arginine (%)	2059	0.86350	0.00900	0.86631	0.03425	0.01509	21	-0.08	121.00	0
121	Arginine (%)	0652	0.86500	0.05000	0.86631	0.03425	0.01509	21	-0.04	121.00	0
121	Arginine (%)	0644	0.86750	0.00500	0.86631	0.03425	0.01509	21	0.03	121.00	0
121	Arginine (%)	0571	0.86950	0.00100	0.86631	0.03425	0.01509	21	0.09	121.00	0
121	Arginine (%)	0297	0.88400	0.01600	0.86631	0.03425	0.01509	21	0.52	121.01	0
121	Arginine (%)	0918	0.88500	0.01800	0.86631	0.03425	0.01509	21	0.55	121.00	0
121	Arginine (%)	0868	0.89185	0.00910	0.86631	0.03425	0.01509	21	0.75	121.00	0
121	Arginine (%)	2004	0.89200	0.01400	0.86631	0.03425	0.01509	21	0.75	121.00	0
121	Arginine (%)	2023	0.89500	0.01000	0.86631	0.03425	0.01509	21	0.84	121.02	0
121	Arginine (%)	0676	0.91500	0.03000	0.86631	0.03425	0.01509	21	1.42	121.05	0
121	Arginine (%)	2022	0.95000	0.06000	0.86631	0.03425	0.01509	21	2.44	121.01	0
121	Arginine (%)	2022	1.0350	0.01000	0.86631	0.03425	0.01509	21	4.93	121.00	0
121	Arginine (%)	0008	0.73450	0.10500	0.86631	0.03425	0.01509	21	-3.85	121.05	1
122	Aspartic (%)	0948	0.92850	0.01900	1.1141	0.02737	0.01359	21	-6.78	122.01	0
122	Aspartic (%)	2022	1.0500	0.02000	1.1141	0.02737	0.01359	21	-2.34	122.00	0
122	Aspartic (%)	0297	1.0705	0.00700	1.1141	0.02737	0.01359	21	-1.59	122.01	0
122	Aspartic (%)	0918	1.0830	0.02200	1.1141	0.02737	0.01359	21	-1.14	122.00	0
122	Aspartic (%)	0504	1.0950	0.01000	1.1141	0.02737	0.01359	21	-0.70	122.00	0
122	Aspartic (%)	2004	1.1000	0.00000	1.1141	0.02737	0.01359	21	-0.52	122.00	0
122	Aspartic (%)	2023	1.1000	0.02000	1.1141	0.02737	0.01359	21	-0.52	122.02	0
122	Aspartic (%)	2059	1.1010	0.00200	1.1141	0.02737	0.01359	21	-0.48	122.00	0
122	Aspartic (%)	0968	1.1055	0.00100	1.1141	0.02737	0.01359	21	-0.32	122.00	0
122	Aspartic (%)	0684	1.1075	0.01900	1.1141	0.02737	0.01359	21	-0.24	122.00	0
122	Aspartic (%)	0619	1.1170	0.02200	1.1141	0.02737	0.01359	21	0.10	122.00	0
122	Aspartic (%)	0571	1.1195	0.00300	1.1141	0.02737	0.01359	21	0.20	122.00	0
122	Aspartic (%)	0652	1.1200	0.02000	1.1141	0.02737	0.01359	21	0.21	122.00	0
122	Aspartic (%)	0676	1.1250	0.07000	1.1141	0.02737	0.01359	21	0.40	122.05	0
122	Aspartic (%)	0098	1.1320	0.00200	1.1141	0.02737	0.01359	21	0.65	122.02	0
122	Aspartic (%)	0872	1.1344	0.00480	1.1141	0.02737	0.01359	21	0.74	122.00	0
122	Aspartic (%)	0675	1.1350	0.01000	1.1141	0.02737	0.01359	21	0.76	122.00	0
122	Aspartic (%)	0227	1.1400	0.00000	1.1141	0.02737	0.01359	21	0.94	122.00	0
122	Aspartic (%)	0644	1.1420	0.00400	1.1141	0.02737	0.01359	21	1.02	122.00	0
122	Aspartic (%)	2022	1.1500	0.02000	1.1141	0.02737	0.01359	21	1.31	122.01	0
122	Aspartic (%)	0868	1.1728	0.00950	1.1141	0.02737	0.01359	21	2.14	122.00	0
122	Aspartic (%)	0008	1.0690	0.12400	1.1141	0.02737	0.01359	21	-1.65	122.05	1
124	Cysteine/Cystine (%)	0008	0.15800	0.00800	0.29084	0.02170	0.00915	19	-6.12	124.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
124	Cysteine/Cystine (%)	0644	0.26800	0.00000	0.29084	0.02170	0.00915	19	-1.05	124.00	0
124	Cysteine/Cystine (%)	0297	0.26800	0.00200	0.29084	0.02170	0.00915	19	-1.05	124.01	0
124	Cysteine/Cystine (%)	2022	0.27000	0.02000	0.29084	0.02170	0.00915	19	-0.96	124.00	0
124	Cysteine/Cystine (%)	0868	0.27470	0.00080	0.29084	0.02170	0.00915	19	-0.74	124.00	0
124	Cysteine/Cystine (%)	0227	0.28000	0.02000	0.29084	0.02170	0.00915	19	-0.50	124.00	0
124	Cysteine/Cystine (%)	0675	0.28500	0.01000	0.29084	0.02170	0.00915	19	-0.27	124.00	0
124	Cysteine/Cystine (%)	0872	0.28670	0.00180	0.29084	0.02170	0.00915	19	-0.19	124.00	0
124	Cysteine/Cystine (%)	0968	0.28950	0.00300	0.29084	0.02170	0.00915	19	-0.06	124.00	0
124	Cysteine/Cystine (%)	2004	0.28950	0.01500	0.29084	0.02170	0.00915	19	-0.06	124.00	0
124	Cysteine/Cystine (%)	0504	0.29000	0.00000	0.29084	0.02170	0.00915	19	-0.04	124.00	0
124	Cysteine/Cystine (%)	0571	0.29300	0.00200	0.29084	0.02170	0.00915	19	0.10	124.00	0
124	Cysteine/Cystine (%)	0684	0.29350	0.01500	0.29084	0.02170	0.00915	19	0.12	124.00	0
124	Cysteine/Cystine (%)	0652	0.29500	0.01000	0.29084	0.02170	0.00915	19	0.19	124.00	0
124	Cysteine/Cystine (%)	2059	0.30900	0.00400	0.29084	0.02170	0.00915	19	0.84	124.00	0
124	Cysteine/Cystine (%)	0098	0.31000	0.00400	0.29084	0.02170	0.00915	19	0.88	124.02	0
124	Cysteine/Cystine (%)	0619	0.32085	0.00830	0.29084	0.02170	0.00915	19	1.38	124.00	0
124	Cysteine/Cystine (%)	0676	0.32500	0.01000	0.29084	0.02170	0.00915	19	1.57	124.02	0
124	Cysteine/Cystine (%)	2023	0.33000	0.04000	0.29084	0.02170	0.00915	19	1.80	124.02	0
124	Cysteine/Cystine (%)	0918	0.29900	0.05640	0.29084	0.02170	0.00915	19	0.38	124.00	1
125	Glutamic (%)	0504	2.2650	0.11000	2.5984	0.07750	0.04456	21	-4.30	125.00	0
125	Glutamic (%)	0948	2.4575	0.04100	2.5984	0.07750	0.04456	21	-1.82	125.01	0
125	Glutamic (%)	2059	2.5280	0.02600	2.5984	0.07750	0.04456	21	-0.91	125.00	0
125	Glutamic (%)	0619	2.5430	0.07490	2.5984	0.07750	0.04456	21	-0.72	125.00	0
125	Glutamic (%)	0098	2.5445	0.05500	2.5984	0.07750	0.04456	21	-0.70	125.02	0
125	Glutamic (%)	2023	2.5450	0.03000	2.5984	0.07750	0.04456	21	-0.69	125.02	0
125	Glutamic (%)	0571	2.5630	0.00800	2.5984	0.07750	0.04456	21	-0.46	125.00	0
125	Glutamic (%)	0227	2.5700	0.02000	2.5984	0.07750	0.04456	21	-0.37	125.00	0
125	Glutamic (%)	0297	2.5740	0.05000	2.5984	0.07750	0.04456	21	-0.32	125.01	0
125	Glutamic (%)	0968	2.5800	0.01400	2.5984	0.07750	0.04456	21	-0.24	125.00	0
125	Glutamic (%)	2022	2.5950	0.09000	2.5984	0.07750	0.04456	21	-0.04	125.00	0
125	Glutamic (%)	0918	2.5995	0.08500	2.5984	0.07750	0.04456	21	0.01	125.00	0
125	Glutamic (%)	2004	2.6050	0.01000	2.5984	0.07750	0.04456	21	0.08	125.00	0
125	Glutamic (%)	0872	2.6173	0.02210	2.5984	0.07750	0.04456	21	0.24	125.00	0
125	Glutamic (%)	0644	2.6195	0.00100	2.5984	0.07750	0.04456	21	0.27	125.00	0
125	Glutamic (%)	2022	2.6550	0.05000	2.5984	0.07750	0.04456	21	0.73	125.01	0
125	Glutamic (%)	0652	2.6650	0.01000	2.5984	0.07750	0.04456	21	0.86	125.00	0
125	Glutamic (%)	0676	2.6800	0.18000	2.5984	0.07750	0.04456	21	1.05	125.05	0
125	Glutamic (%)	0868	2.7030	0.03480	2.5984	0.07750	0.04456	21	1.35	125.00	0
125	Glutamic (%)	0684	2.7250	0.01400	2.5984	0.07750	0.04456	21	1.63	125.00	0
125	Glutamic (%)	0675	2.9450	0.01000	2.5984	0.07750	0.04456	21	4.47	125.00	0
125	Glutamic (%)	0008	2.5300	0.34800	2.5984	0.07750	0.04456	21	-0.88	125.05	1
126	Glycine (%)	0948	0.57700	0.00600	0.70734	0.01858	0.00968	21	-7.01	126.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
126	Glycine (%)	0297	0.65150	0.01100	0.70734	0.01858	0.00968	21	-3.01	126.01	0
126	Glycine (%)	2022	0.68000	0.02000	0.70734	0.01858	0.00968	21	-1.47	126.00	0
126	Glycine (%)	0644	0.68200	0.00600	0.70734	0.01858	0.00968	21	-1.36	126.00	0
126	Glycine (%)	0504	0.69000	0.00000	0.70734	0.01858	0.00968	21	-0.93	126.00	0
126	Glycine (%)	0918	0.69750	0.01500	0.70734	0.01858	0.00968	21	-0.53	126.00	0
126	Glycine (%)	0968	0.69800	0.01000	0.70734	0.01858	0.00968	21	-0.50	126.00	0
126	Glycine (%)	0652	0.70000	0.00000	0.70734	0.01858	0.00968	21	-0.39	126.00	0
126	Glycine (%)	0872	0.70630	0.00360	0.70734	0.01858	0.00968	21	-0.06	126.00	0
126	Glycine (%)	2059	0.70750	0.00300	0.70734	0.01858	0.00968	21	0.01	126.00	0
126	Glycine (%)	0571	0.70900	0.00400	0.70734	0.01858	0.00968	21	0.09	126.00	0
126	Glycine (%)	0098	0.71000	0.00200	0.70734	0.01858	0.00968	21	0.14	126.02	0
126	Glycine (%)	2022	0.71500	0.01000	0.70734	0.01858	0.00968	21	0.41	126.01	0
126	Glycine (%)	2023	0.71500	0.01000	0.70734	0.01858	0.00968	21	0.41	126.02	0
126	Glycine (%)	0684	0.71550	0.01300	0.70734	0.01858	0.00968	21	0.44	126.00	0
126	Glycine (%)	0227	0.72000	0.02000	0.70734	0.01858	0.00968	21	0.68	126.00	0
126	Glycine (%)	2004	0.72050	0.00900	0.70734	0.01858	0.00968	21	0.71	126.00	0
126	Glycine (%)	0619	0.72140	0.01540	0.70734	0.01858	0.00968	21	0.76	126.00	0
126	Glycine (%)	0868	0.72555	0.00530	0.70734	0.01858	0.00968	21	0.98	126.00	0
126	Glycine (%)	0675	0.74500	0.01000	0.70734	0.01858	0.00968	21	2.03	126.00	0
126	Glycine (%)	0676	0.74500	0.03000	0.70734	0.01858	0.00968	21	2.03	126.05	0
126	Glycine (%)	0008	0.67450	0.08700	0.70734	0.01858	0.00968	21	-1.77	126.05	1
127	Histidine (%)	0297	0.32200	0.00400	0.40265	0.01159	0.00599	19	-6.96	127.01	0
127	Histidine (%)	2022	0.35500	0.01000	0.40265	0.01159	0.00599	19	-4.11	127.00	0
127	Histidine (%)	2023	0.38000	0.00000	0.40265	0.01159	0.00599	19	-1.95	127.02	0
127	Histidine (%)	0675	0.39500	0.01000	0.40265	0.01159	0.00599	19	-0.66	127.00	0
127	Histidine (%)	0684	0.39500	0.00200	0.40265	0.01159	0.00599	19	-0.66	127.00	0
127	Histidine (%)	0644	0.39900	0.00200	0.40265	0.01159	0.00599	19	-0.31	127.00	0
127	Histidine (%)	0227	0.40000	0.00000	0.40265	0.01159	0.00599	19	-0.23	127.00	0
127	Histidine (%)	0504	0.40000	0.02000	0.40265	0.01159	0.00599	19	-0.23	127.00	0
127	Histidine (%)	2059	0.40050	0.00300	0.40265	0.01159	0.00599	19	-0.19	127.00	0
127	Histidine (%)	0968	0.40200	0.00200	0.40265	0.01159	0.00599	19	-0.06	127.00	0
127	Histidine (%)	0571	0.40250	0.01300	0.40265	0.01159	0.00599	19	-0.01	127.00	0
127	Histidine (%)	0652	0.40500	0.01000	0.40265	0.01159	0.00599	19	0.20	127.00	0
127	Histidine (%)	0676	0.40500	0.01000	0.40265	0.01159	0.00599	19	0.20	127.05	0
127	Histidine (%)	0872	0.40780	0.00060	0.40265	0.01159	0.00599	19	0.44	127.00	0
127	Histidine (%)	2004	0.41100	0.00200	0.40265	0.01159	0.00599	19	0.72	127.00	0
127	Histidine (%)	0918	0.41550	0.00700	0.40265	0.01159	0.00599	19	1.11	127.00	0
127	Histidine (%)	0868	0.42115	0.00370	0.40265	0.01159	0.00599	19	1.60	127.00	0
127	Histidine (%)	0098	0.42500	0.01000	0.40265	0.01159	0.00599	19	1.93	127.02	0
127	Histidine (%)	0619	0.44220	0.00460	0.40265	0.01159	0.00599	19	3.41	127.00	0
127	Histidine (%)	0008	0.37100	0.03000	0.40265	0.01159	0.00599	19	-2.73	127.05	1
128	Isoleucine (%)	0948	0.43700	0.00000	0.50783	0.02238	0.01219	21	-3.16	128.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
128	Isoleucine (%)	2022	0.47000	0.02000	0.50783	0.02238	0.01219	21	-1.69	128.00	0
128	Isoleucine (%)	2023	0.47500	0.01000	0.50783	0.02238	0.01219	21	-1.47	128.02	0
128	Isoleucine (%)	0675	0.48500	0.01000	0.50783	0.02238	0.01219	21	-1.02	128.00	0
128	Isoleucine (%)	0227	0.49000	0.02000	0.50783	0.02238	0.01219	21	-0.80	128.00	0
128	Isoleucine (%)	0684	0.49450	0.00700	0.50783	0.02238	0.01219	21	-0.60	128.00	0
128	Isoleucine (%)	0619	0.49500	0.00400	0.50783	0.02238	0.01219	21	-0.57	128.00	0
128	Isoleucine (%)	0868	0.50635	0.00710	0.50783	0.02238	0.01219	21	-0.07	128.00	0
128	Isoleucine (%)	0297	0.50750	0.04500	0.50783	0.02238	0.01219	21	-0.01	128.01	0
128	Isoleucine (%)	0644	0.50900	0.00400	0.50783	0.02238	0.01219	21	0.05	128.00	0
128	Isoleucine (%)	0968	0.50900	0.00200	0.50783	0.02238	0.01219	21	0.05	128.00	0
128	Isoleucine (%)	2059	0.51000	0.00200	0.50783	0.02238	0.01219	21	0.10	128.00	0
128	Isoleucine (%)	0571	0.51250	0.01100	0.50783	0.02238	0.01219	21	0.21	128.00	0
128	Isoleucine (%)	0872	0.51255	0.00390	0.50783	0.02238	0.01219	21	0.21	128.00	0
128	Isoleucine (%)	0918	0.51400	0.00400	0.50783	0.02238	0.01219	21	0.28	128.00	0
128	Isoleucine (%)	0652	0.51500	0.01000	0.50783	0.02238	0.01219	21	0.32	128.00	0
128	Isoleucine (%)	0676	0.52000	0.02000	0.50783	0.02238	0.01219	21	0.54	128.05	0
128	Isoleucine (%)	2004	0.53000	0.01400	0.50783	0.02238	0.01219	21	0.99	128.00	0
128	Isoleucine (%)	0098	0.54200	0.00200	0.50783	0.02238	0.01219	21	1.53	128.02	0
128	Isoleucine (%)	0504	0.54500	0.01000	0.50783	0.02238	0.01219	21	1.66	128.00	0
128	Isoleucine (%)	2022	0.60500	0.05000	0.50783	0.02238	0.01219	21	4.34	128.01	0
128	Isoleucine (%)	0008	0.43250	0.06300	0.50783	0.02238	0.01219	21	-3.37	128.05	1
129	Leucine (%)	0948	1.0715	0.02500	1.2075	0.02841	0.02006	20	-4.79	129.01	0
129	Leucine (%)	2022	1.1050	0.05000	1.2075	0.02841	0.02006	20	-3.61	129.00	0
129	Leucine (%)	0619	1.1745	0.02100	1.2075	0.02841	0.02006	20	-1.16	129.00	0
129	Leucine (%)	0968	1.1825	0.00100	1.2075	0.02841	0.02006	20	-0.88	129.00	0
129	Leucine (%)	0652	1.1900	0.02000	1.2075	0.02841	0.02006	20	-0.62	129.00	0
129	Leucine (%)	2059	1.1965	0.00900	1.2075	0.02841	0.02006	20	-0.39	129.00	0
129	Leucine (%)	0297	1.1975	0.02100	1.2075	0.02841	0.02006	20	-0.35	129.01	0
129	Leucine (%)	0684	1.1995	0.00700	1.2075	0.02841	0.02006	20	-0.28	129.00	0
129	Leucine (%)	0227	1.2050	0.03000	1.2075	0.02841	0.02006	20	-0.09	129.00	0
129	Leucine (%)	0872	1.2053	0.01590	1.2075	0.02841	0.02006	20	-0.08	129.00	0
129	Leucine (%)	0868	1.2080	0.01320	1.2075	0.02841	0.02006	20	0.02	129.00	0
129	Leucine (%)	0644	1.2080	0.00600	1.2075	0.02841	0.02006	20	0.02	129.00	0
129	Leucine (%)	0918	1.2085	0.07100	1.2075	0.02841	0.02006	20	0.03	129.00	0
129	Leucine (%)	0571	1.2170	0.00400	1.2075	0.02841	0.02006	20	0.33	129.00	0
129	Leucine (%)	2023	1.2300	0.00000	1.2075	0.02841	0.02006	20	0.79	129.02	0
129	Leucine (%)	0504	1.2350	0.01000	1.2075	0.02841	0.02006	20	0.97	129.00	0
129	Leucine (%)	2004	1.2350	0.01000	1.2075	0.02841	0.02006	20	0.97	129.00	0
129	Leucine (%)	0676	1.2350	0.07000	1.2075	0.02841	0.02006	20	0.97	129.05	0
129	Leucine (%)	0098	1.2435	0.00700	1.2075	0.02841	0.02006	20	1.27	129.02	0
129	Leucine (%)	0675	1.3150	0.01000	1.2075	0.02841	0.02006	20	3.78	129.00	0
129	Leucine (%)	0008	1.1255	0.17300	1.2075	0.02841	0.02006	20	-2.89	129.05	1

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
130	L-Lysine (%)	0910	0.00000	0.00000	0.54948	0.03781	0.01205	22	-14.53	130.00	0
130	L-Lysine (%)	0948	0.44500	0.05600	0.54948	0.03781	0.01205	22	-2.76	130.01	0
130	L-Lysine (%)	0297	0.50100	0.00600	0.54948	0.03781	0.01205	22	-1.28	130.01	0
130	L-Lysine (%)	0684	0.51000	0.00400	0.54948	0.03781	0.01205	22	-1.04	130.00	0
130	L-Lysine (%)	0619	0.52400	0.00400	0.54948	0.03781	0.01205	22	-0.67	130.00	0
130	L-Lysine (%)	0723	0.52400	0.00000	0.54948	0.03781	0.01205	22	-0.67	130.05	0
130	L-Lysine (%)	0652	0.53500	0.03000	0.54948	0.03781	0.01205	22	-0.38	130.00	0
130	L-Lysine (%)	0571	0.53850	0.01300	0.54948	0.03781	0.01205	22	-0.29	130.00	0
130	L-Lysine (%)	2059	0.54350	0.00100	0.54948	0.03781	0.01205	22	-0.16	130.00	0
130	L-Lysine (%)	0675	0.54500	0.01000	0.54948	0.03781	0.01205	22	-0.12	130.00	0
130	L-Lysine (%)	2022	0.54500	0.03000	0.54948	0.03781	0.01205	22	-0.12	130.00	0
130	L-Lysine (%)	0872	0.54615	0.00050	0.54948	0.03781	0.01205	22	-0.09	130.00	0
130	L-Lysine (%)	0644	0.54800	0.00200	0.54948	0.03781	0.01205	22	-0.04	130.00	0
130	L-Lysine (%)	0676	0.55500	0.03000	0.54948	0.03781	0.01205	22	0.15	130.05	0
130	L-Lysine (%)	0918	0.56600	0.04000	0.54948	0.03781	0.01205	22	0.44	130.00	0
130	L-Lysine (%)	2004	0.57000	0.00000	0.54948	0.03781	0.01205	22	0.54	130.00	0
130	L-Lysine (%)	0868	0.57655	0.00250	0.54948	0.03781	0.01205	22	0.72	130.00	0
130	L-Lysine (%)	0968	0.58500	0.01000	0.54948	0.03781	0.01205	22	0.94	130.00	0
130	L-Lysine (%)	0098	0.62200	0.01600	0.54948	0.03781	0.01205	22	1.92	130.02	0
130	L-Lysine (%)	0504	0.64000	0.00000	0.54948	0.03781	0.01205	22	2.39	130.00	0
130	L-Lysine (%)	2023	0.68000	0.00000	0.54948	0.03781	0.01205	22	3.45	130.02	0
130	L-Lysine (%)	0227	0.68500	0.01000	0.54948	0.03781	0.01205	22	3.58	130.00	0
130	L-Lysine (%)	0008	0.50950	0.06900	0.54948	0.03781	0.01205	22	-1.06	130.05	1
131	Methionine (%)	2023	0.22500	0.01000	0.25660	0.01949	0.00528	20	-1.62	131.02	0
131	Methionine (%)	2004	0.23350	0.00500	0.25660	0.01949	0.00528	20	-1.19	131.00	0
131	Methionine (%)	2022	0.23500	0.01000	0.25660	0.01949	0.00528	20	-1.11	131.00	0
131	Methionine (%)	0504	0.24000	0.00000	0.25660	0.01949	0.00528	20	-0.85	131.00	0
131	Methionine (%)	0968	0.24150	0.00100	0.25660	0.01949	0.00528	20	-0.77	131.00	0
131	Methionine (%)	2059	0.24500	0.00000	0.25660	0.01949	0.00528	20	-0.59	131.00	0
131	Methionine (%)	0227	0.25000	0.00000	0.25660	0.01949	0.00528	20	-0.34	131.00	0
131	Methionine (%)	0676	0.25000	0.00000	0.25660	0.01949	0.00528	20	-0.34	131.05	0
131	Methionine (%)	0571	0.25050	0.00700	0.25660	0.01949	0.00528	20	-0.31	131.00	0
131	Methionine (%)	0684	0.25250	0.00700	0.25660	0.01949	0.00528	20	-0.21	131.00	0
131	Methionine (%)	0872	0.25625	0.00410	0.25660	0.01949	0.00528	20	-0.02	131.00	0
131	Methionine (%)	0098	0.26000	0.00000	0.25660	0.01949	0.00528	20	0.17	131.02	0
131	Methionine (%)	0676	0.26000	0.02000	0.25660	0.01949	0.00528	20	0.17	131.02	0
131	Methionine (%)	0619	0.26450	0.00280	0.25660	0.01949	0.00528	20	0.41	131.00	0
131	Methionine (%)	0918	0.26950	0.00300	0.25660	0.01949	0.00528	20	0.66	131.00	0
131	Methionine (%)	0652	0.27000	0.02000	0.25660	0.01949	0.00528	20	0.69	131.00	0
131	Methionine (%)	0297	0.27600	0.00000	0.25660	0.01949	0.00528	20	1.00	131.01	0
131	Methionine (%)	0868	0.28095	0.00070	0.25660	0.01949	0.00528	20	1.25	131.00	0
131	Methionine (%)	0644	0.28750	0.00500	0.25660	0.01949	0.00528	20	1.59	131.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
131	Methionine (%)	0675	0.38500	0.01000	0.25660	0.01949	0.00528	20	<b>6.59</b>	131.00	0
131	Methionine (%)	0008	0.14600	0.03200	0.25660	0.01949	0.00528	20	<b>-5.67</b>	131.99	1
132	Phenylalanine (%)	0948	0.58500	0.01400	0.65885	0.02408	0.01298	20	<b>-3.07</b>	132.01	0
132	Phenylalanine (%)	0684	0.62950	0.00300	0.65885	0.02408	0.01298	20	<b>-1.22</b>	132.00	0
132	Phenylalanine (%)	2059	0.63950	0.00100	0.65885	0.02408	0.01298	20	<b>-0.80</b>	132.00	0
132	Phenylalanine (%)	2023	0.64000	0.00000	0.65885	0.02408	0.01298	20	<b>-0.78</b>	132.02	0
132	Phenylalanine (%)	0968	0.64050	0.01100	0.65885	0.02408	0.01298	20	<b>-0.76</b>	132.00	0
132	Phenylalanine (%)	0644	0.64450	0.00100	0.65885	0.02408	0.01298	20	<b>-0.60</b>	132.00	0
132	Phenylalanine (%)	0227	0.64500	0.03000	0.65885	0.02408	0.01298	20	<b>-0.58</b>	132.00	0
132	Phenylalanine (%)	0652	0.64500	0.01000	0.65885	0.02408	0.01298	20	<b>-0.58</b>	132.00	0
132	Phenylalanine (%)	0918	0.64900	0.01000	0.65885	0.02408	0.01298	20	<b>-0.41</b>	132.00	0
132	Phenylalanine (%)	2004	0.65300	0.00400	0.65885	0.02408	0.01298	20	<b>-0.24</b>	132.00	0
132	Phenylalanine (%)	0676	0.66000	0.02000	0.65885	0.02408	0.01298	20	<b>0.05</b>	132.05	0
132	Phenylalanine (%)	0571	0.66350	0.00700	0.65885	0.02408	0.01298	20	<b>0.19</b>	132.00	0
132	Phenylalanine (%)	0872	0.66705	0.00490	0.65885	0.02408	0.01298	20	<b>0.34</b>	132.00	0
132	Phenylalanine (%)	0297	0.66850	0.01900	0.65885	0.02408	0.01298	20	<b>0.40</b>	132.01	0
132	Phenylalanine (%)	0504	0.67500	0.03000	0.65885	0.02408	0.01298	20	<b>0.67</b>	132.00	0
132	Phenylalanine (%)	0868	0.67775	0.00410	0.65885	0.02408	0.01298	20	<b>0.78</b>	132.00	0
132	Phenylalanine (%)	0098	0.67850	0.00300	0.65885	0.02408	0.01298	20	<b>0.82</b>	132.02	0
132	Phenylalanine (%)	0675	0.69500	0.01000	0.65885	0.02408	0.01298	20	<b>1.50</b>	132.00	0
132	Phenylalanine (%)	0619	0.74765	0.01750	0.65885	0.02408	0.01298	20	<b>3.69</b>	132.00	0
132	Phenylalanine (%)	2022	0.80000	0.06000	0.65885	0.02408	0.01298	20	<b>5.86</b>	132.01	0
132	Phenylalanine (%)	0008	0.59550	0.09900	0.65885	0.02408	0.01298	20	<b>-2.63</b>	132.05	1
133	Proline (%)	0675	0.98500	0.01000	1.0378	0.04091	0.02925	15	<b>-1.29</b>	133.00	0
133	Proline (%)	0504	0.99500	0.05000	1.0378	0.04091	0.02925	15	<b>-1.05</b>	133.00	0
133	Proline (%)	2004	0.99750	0.04500	1.0378	0.04091	0.02925	15	<b>-0.99</b>	133.00	0
133	Proline (%)	0652	1.0000	0.02000	1.0378	0.04091	0.02925	15	<b>-0.92</b>	133.00	0
133	Proline (%)	0008	1.0165	0.14500	1.0378	0.04091	0.02925	15	<b>-0.52</b>	133.05	0
133	Proline (%)	0872	1.0175	0.02300	1.0378	0.04091	0.02925	15	<b>-0.50</b>	133.00	0
133	Proline (%)	2059	1.0295	0.01300	1.0378	0.04091	0.02925	15	<b>-0.20</b>	133.00	0
133	Proline (%)	0571	1.0405	0.00300	1.0378	0.04091	0.02925	15	<b>0.07</b>	133.00	0
133	Proline (%)	0227	1.0500	0.04000	1.0378	0.04091	0.02925	15	<b>0.30</b>	133.00	0
133	Proline (%)	0968	1.0505	0.00300	1.0378	0.04091	0.02925	15	<b>0.31</b>	133.00	0
133	Proline (%)	0684	1.0530	0.03000	1.0378	0.04091	0.02925	15	<b>0.37</b>	133.00	0
133	Proline (%)	0644	1.0680	0.01400	1.0378	0.04091	0.02925	15	<b>0.74</b>	133.00	0
133	Proline (%)	0619	1.0688	0.01610	1.0378	0.04091	0.02925	15	<b>0.76</b>	133.00	0
133	Proline (%)	0868	1.0964	0.01670	1.0378	0.04091	0.02925	15	<b>1.43</b>	133.00	0
133	Proline (%)	0676	1.1050	0.01000	1.0378	0.04091	0.02925	15	<b>1.64</b>	133.05	0
133	Proline (%)	0918	1.0850	0.22000	1.0378	0.04091	0.02925	15	<b>1.15</b>	133.00	1
133	Proline (%)	2023	2.2450	0.01000	1.0378	0.04091	0.02925	15	<b>29.51</b>	133.99	2
134	Serine (%)	0504	0.56000	0.00000	0.68210	0.02500	0.01824	21	<b>-4.88</b>	134.00	0
134	Serine (%)	0948	0.59500	0.01000	0.68210	0.02500	0.01824	21	<b>-3.48</b>	134.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
134	Serine (%)	2022	0.64500	0.03000	0.68210	0.02500	0.01824	21	-1.48	134.00	0
134	Serine (%)	0098	0.65800	0.04000	0.68210	0.02500	0.01824	21	-0.96	134.02	0
134	Serine (%)	0918	0.66700	0.01400	0.68210	0.02500	0.01824	21	-0.60	134.00	0
134	Serine (%)	0644	0.67000	0.00400	0.68210	0.02500	0.01824	21	-0.48	134.00	0
134	Serine (%)	2059	0.67350	0.01100	0.68210	0.02500	0.01824	21	-0.34	134.00	0
134	Serine (%)	0968	0.67400	0.00000	0.68210	0.02500	0.01824	21	-0.32	134.00	0
134	Serine (%)	0652	0.67500	0.01000	0.68210	0.02500	0.01824	21	-0.28	134.00	0
134	Serine (%)	0571	0.67950	0.02300	0.68210	0.02500	0.01824	21	-0.10	134.00	0
134	Serine (%)	0297	0.68000	0.00200	0.68210	0.02500	0.01824	21	-0.08	134.01	0
134	Serine (%)	0619	0.68055	0.01070	0.68210	0.02500	0.01824	21	-0.06	134.00	0
134	Serine (%)	0872	0.68655	0.01050	0.68210	0.02500	0.01824	21	0.18	134.00	0
134	Serine (%)	0227	0.69000	0.02000	0.68210	0.02500	0.01824	21	0.32	134.00	0
134	Serine (%)	2022	0.69500	0.09000	0.68210	0.02500	0.01824	21	0.52	134.01	0
134	Serine (%)	0684	0.70100	0.02800	0.68210	0.02500	0.01824	21	0.76	134.00	0
134	Serine (%)	2023	0.70500	0.01000	0.68210	0.02500	0.01824	21	0.92	134.02	0
134	Serine (%)	0868	0.70900	0.00480	0.68210	0.02500	0.01824	21	1.08	134.00	0
134	Serine (%)	2004	0.72350	0.02500	0.68210	0.02500	0.01824	21	1.66	134.00	0
134	Serine (%)	0675	0.74500	0.01000	0.68210	0.02500	0.01824	21	2.52	134.00	0
134	Serine (%)	0676	0.74500	0.03000	0.68210	0.02500	0.01824	21	2.52	134.05	0
134	Serine (%)	0008	0.68700	0.11000	0.68210	0.02500	0.01824	21	0.20	134.05	1
135	Threonine (%)	0948	0.43800	0.01000	0.53573	0.01715	0.01134	21	-5.70	135.01	0
135	Threonine (%)	2004	0.45450	0.00900	0.53573	0.01715	0.01134	21	-4.74	135.00	0
135	Threonine (%)	2022	0.49500	0.01000	0.53573	0.01715	0.01134	21	-2.38	135.00	0
135	Threonine (%)	0918	0.52000	0.01200	0.53573	0.01715	0.01134	21	-0.92	135.00	0
135	Threonine (%)	2059	0.52350	0.00500	0.53573	0.01715	0.01134	21	-0.71	135.00	0
135	Threonine (%)	0652	0.52500	0.01000	0.53573	0.01715	0.01134	21	-0.63	135.00	0
135	Threonine (%)	0968	0.52750	0.00100	0.53573	0.01715	0.01134	21	-0.48	135.00	0
135	Threonine (%)	0571	0.53000	0.00200	0.53573	0.01715	0.01134	21	-0.33	135.00	0
135	Threonine (%)	0619	0.53015	0.01810	0.53573	0.01715	0.01134	21	-0.33	135.00	0
135	Threonine (%)	0684	0.53350	0.01500	0.53573	0.01715	0.01134	21	-0.13	135.00	0
135	Threonine (%)	0504	0.53500	0.03000	0.53573	0.01715	0.01134	21	-0.04	135.00	0
135	Threonine (%)	0297	0.53600	0.00200	0.53573	0.01715	0.01134	21	0.02	135.01	0
135	Threonine (%)	0872	0.53995	0.00310	0.53573	0.01715	0.01134	21	0.25	135.00	0
135	Threonine (%)	2023	0.54000	0.00000	0.53573	0.01715	0.01134	21	0.25	135.02	0
135	Threonine (%)	0227	0.54500	0.01000	0.53573	0.01715	0.01134	21	0.54	135.00	0
135	Threonine (%)	0675	0.54500	0.01000	0.53573	0.01715	0.01134	21	0.54	135.00	0
135	Threonine (%)	0644	0.55250	0.00100	0.53573	0.01715	0.01134	21	0.98	135.00	0
135	Threonine (%)	0868	0.55935	0.00490	0.53573	0.01715	0.01134	21	1.38	135.00	0
135	Threonine (%)	0676	0.56500	0.03000	0.53573	0.01715	0.01134	21	1.71	135.05	0
135	Threonine (%)	2022	0.58000	0.02000	0.53573	0.01715	0.01134	21	2.58	135.01	0
135	Threonine (%)	0098	0.58050	0.03500	0.53573	0.01715	0.01134	21	2.61	135.02	0
135	Threonine (%)	0008	0.51100	0.07600	0.53573	0.01715	0.01134	21	-1.44	135.05	1

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
136	Tryptophan (%)	0008	0.12800	0.02000	0.16889	0.01342	0.00735	15	-3.05	136.01	0
136	Tryptophan (%)	0941	0.14000	0.00000	0.16889	0.01342	0.00735	15	-2.15	136.00	0
136	Tryptophan (%)	0504	0.15500	0.01000	0.16889	0.01342	0.00735	15	-1.03	136.99	0
136	Tryptophan (%)	2004	0.15950	0.01500	0.16889	0.01342	0.00735	15	-0.70	136.00	0
136	Tryptophan (%)	0868	0.16200	0.00200	0.16889	0.01342	0.00735	15	-0.51	136.01	0
136	Tryptophan (%)	0297	0.16600	0.00200	0.16889	0.01342	0.00735	15	-0.22	136.03	0
136	Tryptophan (%)	0878	0.16650	0.00500	0.16889	0.01342	0.00735	15	-0.18	136.01	0
136	Tryptophan (%)	0872	0.16850	0.00300	0.16889	0.01342	0.00735	15	-0.03	136.03	0
136	Tryptophan (%)	0619	0.17280	0.00620	0.16889	0.01342	0.00735	15	0.29	136.03	0
136	Tryptophan (%)	0098	0.17400	0.00400	0.16889	0.01342	0.00735	15	0.38	136.02	0
136	Tryptophan (%)	2023	0.17500	0.03000	0.16889	0.01342	0.00735	15	0.46	136.02	0
136	Tryptophan (%)	0571	0.17650	0.00100	0.16889	0.01342	0.00735	15	0.57	136.01	0
136	Tryptophan (%)	0644	0.18350	0.00100	0.16889	0.01342	0.00735	15	1.09	136.01	0
136	Tryptophan (%)	0227	0.19500	0.01000	0.16889	0.01342	0.00735	15	1.95	136.00	0
136	Tryptophan (%)	0684	0.21950	0.00100	0.16889	0.01342	0.00735	15	3.77	136.00	0
136	Tryptophan (%)	0918	0.17500	0.04400	0.16889	0.01342	0.00735	15	0.46	136.00	1
137	Tyrosine (%)	0227	0.39500	0.03000	0.44914	0.04005	0.00921	15	-1.35	137.00	0
137	Tyrosine (%)	2023	0.41000	0.00000	0.44914	0.04005	0.00921	15	-0.98	137.02	0
137	Tyrosine (%)	0504	0.41000	0.02000	0.44914	0.04005	0.00921	15	-0.98	137.00	0
137	Tyrosine (%)	0948	0.42200	0.00600	0.44914	0.04005	0.00921	15	-0.68	137.01	0
137	Tyrosine (%)	0968	0.42250	0.00500	0.44914	0.04005	0.00921	15	-0.67	137.00	0
137	Tyrosine (%)	2004	0.43100	0.01000	0.44914	0.04005	0.00921	15	-0.45	137.00	0
137	Tyrosine (%)	0872	0.43845	0.00110	0.44914	0.04005	0.00921	15	-0.27	137.00	0
137	Tyrosine (%)	0098	0.45500	0.01200	0.44914	0.04005	0.00921	15	0.15	137.02	0
137	Tyrosine (%)	0684	0.45600	0.01400	0.44914	0.04005	0.00921	15	0.17	137.00	0
137	Tyrosine (%)	0644	0.45850	0.00300	0.44914	0.04005	0.00921	15	0.23	137.00	0
137	Tyrosine (%)	2059	0.46050	0.00300	0.44914	0.04005	0.00921	15	0.28	137.00	0
137	Tyrosine (%)	0297	0.46800	0.00400	0.44914	0.04005	0.00921	15	0.47	137.01	0
137	Tyrosine (%)	0918	0.49600	0.01000	0.44914	0.04005	0.00921	15	1.17	137.00	0
137	Tyrosine (%)	0676	0.50500	0.01000	0.44914	0.04005	0.00921	15	1.39	137.05	0
137	Tyrosine (%)	0675	0.51500	0.01000	0.44914	0.04005	0.00921	15	1.64	137.00	0
137	Tyrosine (%)	0008	0.33650	0.05700	0.44914	0.04005	0.00921	15	-2.81	137.05	1
138	Valine (%)	0948	0.60800	0.00000	0.71625	0.04591	0.01459	21	-2.36	138.01	0
138	Valine (%)	0684	0.64100	0.00200	0.71625	0.04591	0.01459	21	-1.64	138.00	0
138	Valine (%)	2022	0.66500	0.01000	0.71625	0.04591	0.01459	21	-1.12	138.00	0
138	Valine (%)	0675	0.67500	0.01000	0.71625	0.04591	0.01459	21	-0.90	138.00	0
138	Valine (%)	2023	0.67500	0.01000	0.71625	0.04591	0.01459	21	-0.90	138.02	0
138	Valine (%)	0619	0.69050	0.00700	0.71625	0.04591	0.01459	21	-0.56	138.00	0
138	Valine (%)	0968	0.70750	0.01100	0.71625	0.04591	0.01459	21	-0.19	138.00	0
138	Valine (%)	0504	0.71000	0.06000	0.71625	0.04591	0.01459	21	-0.14	138.00	0
138	Valine (%)	0652	0.71000	0.00000	0.71625	0.04591	0.01459	21	-0.14	138.00	0
138	Valine (%)	2059	0.71100	0.00400	0.71625	0.04591	0.01459	21	-0.11	138.00	0



Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
138	Valine (%)	0644	0.71250	0.00700	0.71625	0.04591	0.01459	21	-0.08	138.00	0
138	Valine (%)	0918	0.72000	0.03800	0.71625	0.04591	0.01459	21	0.08	138.00	0
138	Valine (%)	0872	0.72320	0.00140	0.71625	0.04591	0.01459	21	0.15	138.00	0
138	Valine (%)	0571	0.72650	0.01100	0.71625	0.04591	0.01459	21	0.22	138.00	0
138	Valine (%)	0297	0.73200	0.00000	0.71625	0.04591	0.01459	21	0.34	138.01	0
138	Valine (%)	0868	0.74440	0.00900	0.71625	0.04591	0.01459	21	0.61	138.00	0
138	Valine (%)	0227	0.74500	0.03000	0.71625	0.04591	0.01459	21	0.63	138.00	0
138	Valine (%)	0098	0.75200	0.01200	0.71625	0.04591	0.01459	21	0.78	138.02	0
138	Valine (%)	0676	0.78000	0.04000	0.71625	0.04591	0.01459	21	1.39	138.05	0
138	Valine (%)	2004	0.79100	0.02400	0.71625	0.04591	0.01459	21	1.63	138.00	0
138	Valine (%)	2022	0.82000	0.02000	0.71625	0.04591	0.01459	21	2.26	138.01	0
138	Valine (%)	0008	0.63650	0.10100	0.71625	0.04591	0.01459	21	-1.74	138.05	1
139	Taurine (%)	0098	0.00000	0.00000	0.03333	0.05774	0.00000	3	-0.58	139.02	0
139	Taurine (%)	2023	0.00000	0.00000	0.03333	0.05774	0.00000	3	-0.58	139.02	0
139	Taurine (%)	0504	0.10000	0.00000	0.03333	0.05774	0.00000	3	1.15	139.00	0
160	Fructose (%)	0297	0.21650	0.00500	0.29383	0.07932	0.03833	3	-0.97	160.99	0
160	Fructose (%)	0148	0.29000	0.04000	0.29383	0.07932	0.03833	3	-0.05	160.99	0
160	Fructose (%)	0227	0.37500	0.07000	0.29383	0.07932	0.03833	3	1.02	160.99	0
161	Galactose (%)	0297	0.00000	0.00000			0.00000	1			
162	Glucose (%)	0297	0.16850	0.01900	0.36713	0.20866	0.01475	4	-0.95	162.99	0
162	Glucose (%)	0148	0.23500	0.01000	0.36713	0.20866	0.01475	4	-0.63	162.99	0
162	Glucose (%)	0227	0.43500	0.03000	0.36713	0.20866	0.01475	4	0.33	162.99	0
162	Glucose (%)	0629	0.63000	0.00000	0.36713	0.20866	0.01475	4	1.26	162.99	0
163	Lactose (%)	0148	0.00000	0.00000	0.00000	0.00000	0.00000	3	0.00	163.99	0
163	Lactose (%)	0227	0.00000	0.00000	0.00000	0.00000	0.00000	3	0.00	163.99	0
163	Lactose (%)	0297	0.00000	0.00000	0.00000	0.00000	0.00000	3	0.00	163.99	0
164	Maltose (%)	0148	0.00000	0.00000	0.40600	0.46937	0.00500	4	-0.86	164.99	0
164	Maltose (%)	0227	0.00000	0.00000	0.40600	0.46937	0.00500	4	-0.86	164.99	0
164	Maltose (%)	0297	0.78400	0.02000	0.40600	0.46937	0.00500	4	0.81	164.99	0
164	Maltose (%)	0629	0.84000	0.00000	0.40600	0.46937	0.00500	4	0.92	164.99	0
165	Sucrose (%)	0629	1.2800	0.00000	1.6340	0.24687	0.04240	5	-1.43	165.99	0
165	Sucrose (%)	0227	1.5700	0.12000	1.6340	0.24687	0.04240	5	-0.26	165.99	0
165	Sucrose (%)	0148	1.6550	0.01000	1.6340	0.24687	0.04240	5	0.09	165.99	0
165	Sucrose (%)	0199	1.7000	0.04000	1.6340	0.24687	0.04240	5	0.27	165.99	0
165	Sucrose (%)	0297	1.9650	0.04200	1.6340	0.24687	0.04240	5	1.34	165.99	0
166	Raffinose (%)	0297	0.74300	0.00000			0.00000	1			
167	Stachyose (%)	0297	0.33850	0.01100			0.01100	1			
210	Linoleic Acid (%)	0008	2.5700	0.02000			0.02000	1			
400	Water activity (Units)	0942	0.41000	0.02000	0.44551	0.04818	0.00573	4	-0.74	400.01	0
400	Water activity (Units)	0589	0.41650	0.00100	0.44551	0.04818	0.00573	4	-0.60	400.01	0
400	Water activity (Units)	0001	0.44050	0.00100	0.44551	0.04818	0.00573	4	-0.10	400.01	0
400	Water activity (Units)	0552	0.51505	0.00090	0.44551	0.04818	0.00573	4	1.44	400.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
516	Arsenic, total (mg / kg (ppm))	0563	0.10320	0.00980	0.20261	0.02916	0.01034	12	-3.41	516.00	0
516	Arsenic, total (mg / kg (ppm))	2018	0.17000	0.02000	0.20261	0.02916	0.01034	12	-1.12	516.00	0
516	Arsenic, total (mg / kg (ppm))	0199	0.18000	0.02000	0.20261	0.02916	0.01034	12	-0.78	516.53	0
516	Arsenic, total (mg / kg (ppm))	0964	0.18115	0.01930	0.20261	0.02916	0.01034	12	-0.74	516.43	0
516	Arsenic, total (mg / kg (ppm))	0168	0.19500	0.01000	0.20261	0.02916	0.01034	12	-0.26	516.53	0
516	Arsenic, total (mg / kg (ppm))	0553	0.20500	0.01000	0.20261	0.02916	0.01034	12	0.08	516.53	0
516	Arsenic, total (mg / kg (ppm))	0171	0.21000	0.00000	0.20261	0.02916	0.01034	12	0.25	516.00	0
516	Arsenic, total (mg / kg (ppm))	0098	0.21150	0.01500	0.20261	0.02916	0.01034	12	0.30	516.52	0
516	Arsenic, total (mg / kg (ppm))	0227	0.21300	0.01000	0.20261	0.02916	0.01034	12	0.36	516.53	0
516	Arsenic, total (mg / kg (ppm))	2058	0.21500	0.01000	0.20261	0.02916	0.01034	12	0.42	516.53	0
516	Arsenic, total (mg / kg (ppm))	0003	0.24000	0.00000	0.20261	0.02916	0.01034	12	1.28	516.43	0
516	Arsenic, total (mg / kg (ppm))	0939	0.33000	0.00000	0.20261	0.02916	0.01034	12	4.37	516.00	0
518	Cadmium (mg / kg (ppm))	0425	0.00000	0.00000	0.07611	0.02004	0.00357	14	-3.80	518.43	0
518	Cadmium (mg / kg (ppm))	0003	0.04000	0.00000	0.07611	0.02004	0.00357	14	-1.80	518.43	0
518	Cadmium (mg / kg (ppm))	2058	0.06000	0.00000	0.07611	0.02004	0.00357	14	-0.80	518.53	0
518	Cadmium (mg / kg (ppm))	0619	0.06250	0.01100	0.07611	0.02004	0.00357	14	-0.68	518.41	0
518	Cadmium (mg / kg (ppm))	0968	0.06500	0.01000	0.07611	0.02004	0.00357	14	-0.55	518.41	0
518	Cadmium (mg / kg (ppm))	0011	0.06875	0.00850	0.07611	0.02004	0.00357	14	-0.37	518.41	0
518	Cadmium (mg / kg (ppm))	2018	0.07500	0.01000	0.07611	0.02004	0.00357	14	-0.06	518.34	0
518	Cadmium (mg / kg (ppm))	0199	0.08000	0.00000	0.07611	0.02004	0.00357	14	0.19	518.53	0
518	Cadmium (mg / kg (ppm))	0168	0.08850	0.00100	0.07611	0.02004	0.00357	14	0.62	518.53	0
518	Cadmium (mg / kg (ppm))	0227	0.08900	0.00200	0.07611	0.02004	0.00357	14	0.64	518.53	0
518	Cadmium (mg / kg (ppm))	0939	0.09000	0.00000	0.07611	0.02004	0.00357	14	0.69	518.34	0
518	Cadmium (mg / kg (ppm))	0098	0.09100	0.00000	0.07611	0.02004	0.00357	14	0.74	518.52	0
518	Cadmium (mg / kg (ppm))	0171	0.09250	0.00300	0.07611	0.02004	0.00357	14	0.82	518.41	0
518	Cadmium (mg / kg (ppm))	0563	0.12225	0.00450	0.07611	0.02004	0.00357	14	2.30	518.34	0
518	Cadmium (mg / kg (ppm))	0553	0.09360	0.02280	0.07611	0.02004	0.00357	14	0.87	518.53	1
520	Chromium (mg / kg (ppm))	0619	5.0200	0.40000	10.645	4.7871	0.90410	10	-1.17	520.41	0
520	Chromium (mg / kg (ppm))	0171	5.7750	0.51000	10.645	4.7871	0.90410	10	-1.02	520.41	0
520	Chromium (mg / kg (ppm))	0563	6.1400	0.14000	10.645	4.7871	0.90410	10	-0.94	520.31	0
520	Chromium (mg / kg (ppm))	0011	7.9935	0.36000	10.645	4.7871	0.90410	10	-0.55	520.41	0
520	Chromium (mg / kg (ppm))	0510	11.850	0.16000	10.645	4.7871	0.90410	10	0.25	520.43	0
520	Chromium (mg / kg (ppm))	0968	11.854	0.86400	10.645	4.7871	0.90410	10	0.25	520.99	0
520	Chromium (mg / kg (ppm))	0553	12.250	2.1000	10.645	4.7871	0.90410	10	0.34	520.53	0
520	Chromium (mg / kg (ppm))	2023	13.600	1.4000	10.645	4.7871	0.90410	10	0.62	520.53	0
520	Chromium (mg / kg (ppm))	0964	14.123	0.18700	10.645	4.7871	0.90410	10	0.73	520.43	0
520	Chromium (mg / kg (ppm))	0003	18.060	2.9200	10.645	4.7871	0.90410	10	1.55	520.43	0
526	Lead (mg / kg (ppm))	0619	0.00000	0.00000	0.12263	0.07269	0.04248	15	-1.69	526.41	0
526	Lead (mg / kg (ppm))	0425	0.00000	0.00000	0.12263	0.07269	0.04248	15	-1.69	526.43	0
526	Lead (mg / kg (ppm))	2018	0.06000	0.04000	0.12263	0.07269	0.04248	15	-0.86	526.34	0
526	Lead (mg / kg (ppm))	0921	0.07460	0.08460	0.12263	0.07269	0.04248	15	-0.66	526.31	0
526	Lead (mg / kg (ppm))	0003	0.08000	0.16000	0.12263	0.07269	0.04248	15	-0.59	526.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO CS Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rob SD	R-bar	# Labs			
526	Lead (mg / kg (ppm))	2058	0.10000	0.04000	0.12263	0.07269	0.04248	15	-0.31	526.53	0
526	Lead (mg / kg (ppm))	0171	0.11050	0.06100	0.12263	0.07269	0.04248	15	-0.17	526.41	0
526	Lead (mg / kg (ppm))	0199	0.13000	0.00000	0.12263	0.07269	0.04248	15	0.10	526.53	0
526	Lead (mg / kg (ppm))	0964	0.14860	0.04260	0.12263	0.07269	0.04248	15	0.36	526.43	0
526	Lead (mg / kg (ppm))	0939	0.15000	0.00000	0.12263	0.07269	0.04248	15	0.38	526.34	0
526	Lead (mg / kg (ppm))	0553	0.15800	0.00800	0.12263	0.07269	0.04248	15	0.49	526.53	0
526	Lead (mg / kg (ppm))	0227	0.16750	0.02300	0.12263	0.07269	0.04248	15	0.62	526.53	0
526	Lead (mg / kg (ppm))	0168	0.19500	0.01000	0.12263	0.07269	0.04248	15	1.00	526.53	0
526	Lead (mg / kg (ppm))	0098	0.19600	0.09800	0.12263	0.07269	0.04248	15	1.01	526.52	0
526	Lead (mg / kg (ppm))	0563	0.62500	0.07000	0.12263	0.07269	0.04248	15	6.91	526.34	0
529	Mercury (µg / kg (ppb))	0425	0.00000	0.00000	10.187	8.8264	6.2533	3	-1.15	529.00	0
529	Mercury (µg / kg (ppb))	0171	15.000	8.0000	10.187	8.8264	6.2533	3	0.55	529.99	0
529	Mercury (µg / kg (ppb))	2018	15.560	10.760	10.187	8.8264	6.2533	3	0.61	529.99	0
539	Nickel (mg / kg (ppm))	0011	6.1605	0.19900	8.5309	3.0075	0.58218	5	-0.79	539.41	0
539	Nickel (mg / kg (ppm))	0553	7.0800	1.2400	8.5309	3.0075	0.58218	5	-0.48	539.53	0
539	Nickel (mg / kg (ppm))	2023	7.5800	0.68000	8.5309	3.0075	0.58218	5	-0.32	539.53	0
539	Nickel (mg / kg (ppm))	0964	8.0738	0.07190	8.5309	3.0075	0.58218	5	-0.15	539.43	0
539	Nickel (mg / kg (ppm))	0003	13.760	0.72000	8.5309	3.0075	0.58218	5	1.74	539.43	0
600	Total Aflatoxin (µg / kg (ppb))	0042	0.00000	0.00000							
600	Total Aflatoxin (µg / kg (ppb))	0689	1.0500	0.50000							
600	Total Aflatoxin (µg / kg (ppb))	0623	1.2500	0.50000							
600	Total Aflatoxin (µg / kg (ppb))	2018	2.5000	0.20000							
600	Total Aflatoxin (µg / kg (ppb))	0589	2.5000	3.0000							
600	Total Aflatoxin (µg / kg (ppb))	0723	2.6600	0.00000							
600	Total Aflatoxin (µg / kg (ppb))	0880	3.8000	1.8000							
610	Deoxynivalenol (µg / kg (ppb))	2018	997.50	5.0000							
620	Total Fumonisin (µg / kg (ppb))	2018	705.00	10.000							
620	Total Fumonisin (µg / kg (ppb))	0227	800.00	400.00							
620	Total Fumonisin (µg / kg (ppb))	0589	925.00	150.00							
640	T-2 (µg / kg (ppb))	0227	111.80	61.600							
650	Zearalenone (µg / kg (ppb))	2018	143.00	4.0000							
650	Zearalenone (µg / kg (ppb))	0227	183.50	31.000							