



Poultry Layer Feed, Medicated

Animal Feed Scheme

Labs Reporting: 198

Test Material Code # 201722

Analyte Proficiency Testing Report

Issue Date : 03/31/2017

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
000	Urea (%)	0278	0.30000	0.00000				1		000.02	0
001	Loss on Drying (%)	0536	7.6000	0.50000	9.3889	0.36304	0.12727	77	-4.93	001.99	0
001	Loss on Drying (%)	0618	8.0250	0.05000	9.3889	0.36304	0.12727	77	-3.76	001.99	0
001	Loss on Drying (%)	0038	8.2900	0.36000	9.3889	0.36304	0.12727	77	-3.03	001.07	0
001	Loss on Drying (%)	0019	8.4300	0.22000	9.3889	0.36304	0.12727	77	-2.64	001.07	0
001	Loss on Drying (%)	0889	8.5850	0.25000	9.3889	0.36304	0.12727	77	-2.21	001.99	0
001	Loss on Drying (%)	0366	8.6500	1.1000	9.3889	0.36304	0.12727	77	-2.04	001.07	0
001	Loss on Drying (%)	2012	8.7150	0.01000	9.3889	0.36304	0.12727	77	-1.86	001.07	0
001	Loss on Drying (%)	2155	8.7850	0.23000	9.3889	0.36304	0.12727	77	-1.66	001.99	0
001	Loss on Drying (%)	0904	8.8000	0.14000	9.3889	0.36304	0.12727	77	-1.62	001.99	0
001	Loss on Drying (%)	0620	8.8750	0.09920	9.3889	0.36304	0.12727	77	-1.42	001.00	0
001	Loss on Drying (%)	0089	8.9500	0.00000	9.3889	0.36304	0.12727	77	-1.21	001.07	0
001	Loss on Drying (%)	0278	8.9850	0.29000	9.3889	0.36304	0.12727	77	-1.11	001.07	0
001	Loss on Drying (%)	0676	9.0300	0.06000	9.3889	0.36304	0.12727	77	-0.99	001.99	0
001	Loss on Drying (%)	0675	9.0550	0.03000	9.3889	0.36304	0.12727	77	-0.92	001.07	0
001	Loss on Drying (%)	0612	9.0900	0.04000	9.3889	0.36304	0.12727	77	-0.82	001.99	0
001	Loss on Drying (%)	0656	9.0950	0.07000	9.3889	0.36304	0.12727	77	-0.81	001.99	0
001	Loss on Drying (%)	0510	9.1000	0.20000	9.3889	0.36304	0.12727	77	-0.80	001.99	0
001	Loss on Drying (%)	2149	9.1500	0.10000	9.3889	0.36304	0.12727	77	-0.66	001.99	0
001	Loss on Drying (%)	0034	9.1650	0.23000	9.3889	0.36304	0.12727	77	-0.62	001.00	0
001	Loss on Drying (%)	0015	9.1800	0.02000	9.3889	0.36304	0.12727	77	-0.58	001.07	0
001	Loss on Drying (%)	0610	9.1850	0.03000	9.3889	0.36304	0.12727	77	-0.56	001.05	0
001	Loss on Drying (%)	0083	9.2000	0.10000	9.3889	0.36304	0.12727	77	-0.52	001.07	0
001	Loss on Drying (%)	0139	9.2100	0.04000	9.3889	0.36304	0.12727	77	-0.49	001.07	0
001	Loss on Drying (%)	0918	9.2250	0.03000	9.3889	0.36304	0.12727	77	-0.45	001.99	0
001	Loss on Drying (%)	0630	9.2300	0.02000	9.3889	0.36304	0.12727	77	-0.44	001.99	0
001	Loss on Drying (%)	0948	9.2300	0.14000	9.3889	0.36304	0.12727	77	-0.44	001.99	0
001	Loss on Drying (%)	0940	9.2800	0.06000	9.3889	0.36304	0.12727	77	-0.30	001.99	0
001	Loss on Drying (%)	0171	9.2900	0.04000	9.3889	0.36304	0.12727	77	-0.27	001.07	0
001	Loss on Drying (%)	0413	9.3000	0.00000	9.3889	0.36304	0.12727	77	-0.24	001.07	0
001	Loss on Drying (%)	2053	9.3000	0.06000	9.3889	0.36304	0.12727	77	-0.24	001.07	0
001	Loss on Drying (%)	0142	9.3050	0.13000	9.3889	0.36304	0.12727	77	-0.23	001.07	0
001	Loss on Drying (%)	0186	9.3300	0.13720	9.3889	0.36304	0.12727	77	-0.16	001.03	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
001	Loss on Drying (%)	0074	9.3400	0.32000	9.3889	0.36304	0.12727	77	-0.13	001.07	0
001	Loss on Drying (%)	0619	9.3450	0.01000	9.3889	0.36304	0.12727	77	-0.12	001.03	0
001	Loss on Drying (%)	0615	9.3550	0.03000	9.3889	0.36304	0.12727	77	-0.09	001.00	0
001	Loss on Drying (%)	0035	9.3750	0.03000	9.3889	0.36304	0.12727	77	-0.04	001.07	0
001	Loss on Drying (%)	0226	9.4300	0.04000	9.3889	0.36304	0.12727	77	0.11	001.07	0
001	Loss on Drying (%)	2009	9.4327	0.00970	9.3889	0.36304	0.12727	77	0.12	001.07	0
001	Loss on Drying (%)	2059	9.4500	0.06000	9.3889	0.36304	0.12727	77	0.17	001.07	0
001	Loss on Drying (%)	0187	9.4550	0.09000	9.3889	0.36304	0.12727	77	0.18	001.07	0
001	Loss on Drying (%)	0893	9.4600	0.02000	9.3889	0.36304	0.12727	77	0.20	001.03	0
001	Loss on Drying (%)	0613	9.4650	0.95000	9.3889	0.36304	0.12727	77	0.21	001.07	0
001	Loss on Drying (%)	0003	9.4700	0.02000	9.3889	0.36304	0.12727	77	0.22	001.07	0
001	Loss on Drying (%)	0629	9.4800	0.00000	9.3889	0.36304	0.12727	77	0.25	001.99	0
001	Loss on Drying (%)	0910	9.5000	0.20000	9.3889	0.36304	0.12727	77	0.31	001.99	0
001	Loss on Drying (%)	0169	9.5050	0.01000	9.3889	0.36304	0.12727	77	0.32	001.00	0
001	Loss on Drying (%)	0581	9.5050	0.11000	9.3889	0.36304	0.12727	77	0.32	001.07	0
001	Loss on Drying (%)	0571	9.5100	0.02000	9.3889	0.36304	0.12727	77	0.33	001.07	0
001	Loss on Drying (%)	0644	9.5200	0.02000	9.3889	0.36304	0.12727	77	0.36	001.99	0
001	Loss on Drying (%)	0843	9.5250	0.27000	9.3889	0.36304	0.12727	77	0.37	001.07	0
001	Loss on Drying (%)	2109	9.5250	0.25000	9.3889	0.36304	0.12727	77	0.37	001.07	0
001	Loss on Drying (%)	0895	9.5450	0.01000	9.3889	0.36304	0.12727	77	0.43	001.03	0
001	Loss on Drying (%)	0903	9.5450	0.01000	9.3889	0.36304	0.12727	77	0.43	001.03	0
001	Loss on Drying (%)	0592	9.5450	0.03000	9.3889	0.36304	0.12727	77	0.43	001.07	0
001	Loss on Drying (%)	0353	9.5500	0.56000	9.3889	0.36304	0.12727	77	0.44	001.07	0
001	Loss on Drying (%)	0590	9.5850	0.13000	9.3889	0.36304	0.12727	77	0.54	001.08	0
001	Loss on Drying (%)	0689	9.6000	0.00000	9.3889	0.36304	0.12727	77	0.58	001.07	0
001	Loss on Drying (%)	0849	9.6125	0.06500	9.3889	0.36304	0.12727	77	0.62	001.07	0
001	Loss on Drying (%)	0049	9.6450	0.05000	9.3889	0.36304	0.12727	77	0.71	001.07	0
001	Loss on Drying (%)	0027	9.6500	0.04000	9.3889	0.36304	0.12727	77	0.72	001.00	0
001	Loss on Drying (%)	0662	9.6534	0.42460	9.3889	0.36304	0.12727	77	0.73	001.99	0
001	Loss on Drying (%)	0037	9.6550	0.05000	9.3889	0.36304	0.12727	77	0.73	001.99	0
001	Loss on Drying (%)	0878	9.6650	0.01000	9.3889	0.36304	0.12727	77	0.76	001.03	0
001	Loss on Drying (%)	0652	9.7000	0.00000	9.3889	0.36304	0.12727	77	0.86	001.08	0
001	Loss on Drying (%)	0098	9.7100	0.10000	9.3889	0.36304	0.12727	77	0.88	001.07	0
001	Loss on Drying (%)	2062	9.7362	0.03110	9.3889	0.36304	0.12727	77	0.96	001.03	0
001	Loss on Drying (%)	0872	9.7650	0.07000	9.3889	0.36304	0.12727	77	1.04	001.07	0
001	Loss on Drying (%)	0504	9.7750	0.01000	9.3889	0.36304	0.12727	77	1.06	001.00	0
001	Loss on Drying (%)	0004	9.8150	0.01000	9.3889	0.36304	0.12727	77	1.17	001.99	0
001	Loss on Drying (%)	0357	9.8455	0.00930	9.3889	0.36304	0.12727	77	1.26	001.99	0
001	Loss on Drying (%)	0693	9.8650	0.33000	9.3889	0.36304	0.12727	77	1.31	001.07	0
001	Loss on Drying (%)	0505	9.8650	0.01000	9.3889	0.36304	0.12727	77	1.31	001.99	0
001	Loss on Drying (%)	0199	9.8800	0.02000	9.3889	0.36304	0.12727	77	1.35	001.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
001	Loss on Drying (%)	0852	9.9400	0.24000	9.3889	0.36304	0.12727	77	1.52	001.99	0
001	Loss on Drying (%)	0183	9.9590	0.12400	9.3889	0.36304	0.12727	77	1.57	001.00	0
001	Loss on Drying (%)	0045	10.100	0.20000	9.3889	0.36304	0.12727	77	1.96	001.07	0
001	Loss on Drying (%)	0529	10.530	0.02000	9.3889	0.36304	0.12727	77	3.14	001.07	0
002	Protein (%)	0539	2.9400	0.10000	18.962	0.29434	101.76	196	-54.43	002.06	0
002	Protein (%)	0183	16.295	0.01000	18.962	0.29434	101.76	196	-9.06	002.01	0
002	Protein (%)	0536	16.410	0.26000	18.962	0.29434	101.76	196	-8.67	002.05	0
002	Protein (%)	0874	17.975	0.91000	18.962	0.29434	101.76	196	-3.35	002.04	0
002	Protein (%)	0643	18.195	0.15000	18.962	0.29434	101.76	196	-2.60	002.99	0
002	Protein (%)	0038	18.240	0.39600	18.962	0.29434	101.76	196	-2.45	002.06	0
002	Protein (%)	0646	18.250	0.06000	18.962	0.29434	101.76	196	-2.42	002.06	0
002	Protein (%)	0018	18.265	0.07000	18.962	0.29434	101.76	196	-2.37	002.06	0
002	Protein (%)	0004	18.435	0.47000	18.962	0.29434	101.76	196	-1.79	002.06	0
002	Protein (%)	0722	18.475	0.15260	18.962	0.29434	101.76	196	-1.65	002.05	0
002	Protein (%)	0852	18.520	0.08000	18.962	0.29434	101.76	196	-1.50	002.05	0
002	Protein (%)	0940	18.530	0.00000	18.962	0.29434	101.76	196	-1.47	002.00	0
002	Protein (%)	2109	18.545	0.11000	18.962	0.29434	101.76	196	-1.42	002.00	0
002	Protein (%)	0653	18.565	0.03000	18.962	0.29434	101.76	196	-1.35	002.01	0
002	Protein (%)	0003	18.577	0.17300	18.962	0.29434	101.76	196	-1.31	002.06	0
002	Protein (%)	2097	18.582	0.16840	18.962	0.29434	101.76	196	-1.29	002.05	0
002	Protein (%)	0675	18.585	0.03000	18.962	0.29434	101.76	196	-1.28	002.05	0
002	Protein (%)	0618	18.590	0.00000	18.962	0.29434	101.76	196	-1.26	002.06	0
002	Protein (%)	0878	18.590	0.08000	18.962	0.29434	101.76	196	-1.26	002.06	0
002	Protein (%)	0652	18.600	0.20000	18.962	0.29434	101.76	196	-1.23	002.01	0
002	Protein (%)	0674	18.600	0.00000	18.962	0.29434	101.76	196	-1.23	002.05	0
002	Protein (%)	0242	18.600	0.02000	18.962	0.29434	101.76	196	-1.23	002.06	0
002	Protein (%)	0563	18.613	0.04540	18.962	0.29434	101.76	196	-1.18	002.08	0
002	Protein (%)	0939	18.615	0.41000	18.962	0.29434	101.76	196	-1.18	002.05	0
002	Protein (%)	2022	18.615	0.01000	18.962	0.29434	101.76	196	-1.18	002.05	0
002	Protein (%)	0098	18.620	0.08000	18.962	0.29434	101.76	196	-1.16	002.08	0
002	Protein (%)	0685	18.620	0.04000	18.962	0.29434	101.76	196	-1.16	002.01	0
002	Protein (%)	0028	18.645	0.01000	18.962	0.29434	101.76	196	-1.08	002.01	0
002	Protein (%)	0619	18.650	0.10000	18.962	0.29434	101.76	196	-1.06	002.05	0
002	Protein (%)	0952	18.665	0.07000	18.962	0.29434	101.76	196	-1.01	002.05	0
002	Protein (%)	2006	18.680	0.88000	18.962	0.29434	101.76	196	-0.96	002.05	0
002	Protein (%)	0968	18.688	0.02100	18.962	0.29434	101.76	196	-0.93	002.01	0
002	Protein (%)	0552	18.690	0.04000	18.962	0.29434	101.76	196	-0.92	002.05	0
002	Protein (%)	0615	18.695	0.21000	18.962	0.29434	101.76	196	-0.91	002.06	0
002	Protein (%)	0910	18.700	0.20000	18.962	0.29434	101.76	196	-0.89	002.01	0
002	Protein (%)	0199	18.705	0.03000	18.962	0.29434	101.76	196	-0.87	002.01	0
002	Protein (%)	0066	18.710	0.04000	18.962	0.29434	101.76	196	-0.86	002.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein (%)	2012	18.710	0.02000	18.962	0.29434	101.76	196	-0.86	002.05	0
002	Protein (%)	0675	18.710	0.02000	18.962	0.29434	101.76	196	-0.86	002.06	0
002	Protein (%)	0849	18.721	0.06200	18.962	0.29434	101.76	196	-0.82	002.05	0
002	Protein (%)	0043	18.735	0.11000	18.962	0.29434	101.76	196	-0.77	002.01	0
002	Protein (%)	0620	18.739	0.11420	18.962	0.29434	101.76	196	-0.76	002.05	0
002	Protein (%)	0693	18.750	0.30000	18.962	0.29434	101.76	196	-0.72	002.06	0
002	Protein (%)	0889	18.755	0.11000	18.962	0.29434	101.76	196	-0.70	002.01	0
002	Protein (%)	0144	18.755	0.01000	18.962	0.29434	101.76	196	-0.70	002.06	0
002	Protein (%)	0885	18.779	0.02320	18.962	0.29434	101.76	196	-0.62	002.05	0
002	Protein (%)	0100	18.780	0.02000	18.962	0.29434	101.76	196	-0.62	002.06	0
002	Protein (%)	2081	18.780	0.94000	18.962	0.29434	101.76	196	-0.62	002.06	0
002	Protein (%)	0354	18.785	0.01000	18.962	0.29434	101.76	196	-0.60	002.05	0
002	Protein (%)	0039	18.789	0.03300	18.962	0.29434	101.76	196	-0.59	002.06	0
002	Protein (%)	2073	18.795	0.11000	18.962	0.29434	101.76	196	-0.57	002.05	0
002	Protein (%)	0229	18.800	0.06000	18.962	0.29434	101.76	196	-0.55	002.06	0
002	Protein (%)	0345	18.800	0.20000	18.962	0.29434	101.76	196	-0.55	002.06	0
002	Protein (%)	0590	18.800	0.20000	18.962	0.29434	101.76	196	-0.55	002.06	0
002	Protein (%)	0589	18.800	0.32000	18.962	0.29434	101.76	196	-0.55	002.06	0
002	Protein (%)	0692	18.800	0.40000	18.962	0.29434	101.76	196	-0.55	002.06	0
002	Protein (%)	0208	18.800	0.40000	18.962	0.29434	101.76	196	-0.55	002.08	0
002	Protein (%)	2053	18.810	0.04000	18.962	0.29434	101.76	196	-0.52	002.06	0
002	Protein (%)	0505	18.810	0.02000	18.962	0.29434	101.76	196	-0.52	002.06	0
002	Protein (%)	0357	18.813	0.00000	18.962	0.29434	101.76	196	-0.51	002.06	0
002	Protein (%)	0861	18.815	0.07000	18.962	0.29434	101.76	196	-0.50	002.06	0
002	Protein (%)	0148	18.820	0.02000	18.962	0.29434	101.76	196	-0.48	002.06	0
002	Protein (%)	0529	18.820	0.34000	18.962	0.29434	101.76	196	-0.48	002.06	0
002	Protein (%)	0598	18.820	0.08000	18.962	0.29434	101.76	196	-0.48	002.06	0
002	Protein (%)	0169	18.835	0.19000	18.962	0.29434	101.76	196	-0.43	002.00	0
002	Protein (%)	0263	18.843	0.05700	18.962	0.29434	101.76	196	-0.40	002.06	0
002	Protein (%)	0032	18.845	0.09100	18.962	0.29434	101.76	196	-0.40	002.06	0
002	Protein (%)	0187	18.850	0.08000	18.962	0.29434	101.76	196	-0.38	002.04	0
002	Protein (%)	0014	18.850	0.30000	18.962	0.29434	101.76	196	-0.38	002.06	0
002	Protein (%)	0019	18.850	0.06000	18.962	0.29434	101.76	196	-0.38	002.06	0
002	Protein (%)	0033	18.850	0.30000	18.962	0.29434	101.76	196	-0.38	002.06	0
002	Protein (%)	0175	18.850	0.10000	18.962	0.29434	101.76	196	-0.38	002.06	0
002	Protein (%)	0294	18.850	0.10000	18.962	0.29434	101.76	196	-0.38	002.06	0
002	Protein (%)	0619	18.850	0.10000	18.962	0.29434	101.76	196	-0.38	002.06	0
002	Protein (%)	2155	18.854	0.73100	18.962	0.29434	101.76	196	-0.37	002.06	0
002	Protein (%)	0205	18.860	0.38000	18.962	0.29434	101.76	196	-0.35	002.06	0
002	Protein (%)	0629	18.860	0.30000	18.962	0.29434	101.76	196	-0.35	002.10	0
002	Protein (%)	2062	18.869	0.00640	18.962	0.29434	101.76	196	-0.31	002.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein (%)	2009	18.869	0.01140	18.962	0.29434	101.76	196	-0.31	002.05	0
002	Protein (%)	0108	18.880	0.24000	18.962	0.29434	101.76	196	-0.28	002.06	0
002	Protein (%)	0941	18.885	0.21000	18.962	0.29434	101.76	196	-0.26	002.05	0
002	Protein (%)	0650	18.885	0.57000	18.962	0.29434	101.76	196	-0.26	002.06	0
002	Protein (%)	0610	18.900	0.00000	18.962	0.29434	101.76	196	-0.21	002.01	0
002	Protein (%)	0689	18.900	0.40000	18.962	0.29434	101.76	196	-0.21	002.05	0
002	Protein (%)	0017	18.900	0.00000	18.962	0.29434	101.76	196	-0.21	002.06	0
002	Protein (%)	0910	18.900	0.20000	18.962	0.29434	101.76	196	-0.21	002.06	0
002	Protein (%)	0956	18.900	0.00000	18.962	0.29434	101.76	196	-0.21	002.06	0
002	Protein (%)	0914	18.910	0.20000	18.962	0.29434	101.76	196	-0.18	002.01	0
002	Protein (%)	0009	18.910	0.02000	18.962	0.29434	101.76	196	-0.18	002.06	0
002	Protein (%)	0592	18.910	0.04000	18.962	0.29434	101.76	196	-0.18	002.06	0
002	Protein (%)	0968	18.912	0.15200	18.962	0.29434	101.76	196	-0.17	002.06	0
002	Protein (%)	2068	18.915	0.05000	18.962	0.29434	101.76	196	-0.16	002.05	0
002	Protein (%)	2152	18.920	0.06000	18.962	0.29434	101.76	196	-0.14	002.05	0
002	Protein (%)	0613	18.920	0.14000	18.962	0.29434	101.76	196	-0.14	002.06	0
002	Protein (%)	0948	18.925	0.01000	18.962	0.29434	101.76	196	-0.12	002.05	0
002	Protein (%)	0074	18.930	0.20000	18.962	0.29434	101.76	196	-0.11	002.06	0
002	Protein (%)	0504	18.935	0.19000	18.962	0.29434	101.76	196	-0.09	002.04	0
002	Protein (%)	0358	18.935	0.13000	18.962	0.29434	101.76	196	-0.09	002.06	0
002	Protein (%)	0026	18.939	0.01800	18.962	0.29434	101.76	196	-0.08	002.06	0
002	Protein (%)	0139	18.940	0.08000	18.962	0.29434	101.76	196	-0.07	002.06	0
002	Protein (%)	0571	18.941	0.02500	18.962	0.29434	101.76	196	-0.07	002.06	0
002	Protein (%)	0265	18.945	0.01000	18.962	0.29434	101.76	196	-0.06	002.06	0
002	Protein (%)	0278	18.950	0.10000	18.962	0.29434	101.76	196	-0.04	002.06	0
002	Protein (%)	0366	18.950	0.30000	18.962	0.29434	101.76	196	-0.04	002.06	0
002	Protein (%)	0298	18.960	0.02000	18.962	0.29434	101.76	196	-0.01	002.06	0
002	Protein (%)	0848	18.960	0.24000	18.962	0.29434	101.76	196	-0.01	002.06	0
002	Protein (%)	0939	18.970	0.44000	18.962	0.29434	101.76	196	0.03	002.06	0
002	Protein (%)	0903	18.975	0.01000	18.962	0.29434	101.76	196	0.05	002.05	0
002	Protein (%)	0895	18.980	0.02000	18.962	0.29434	101.76	196	0.06	002.05	0
002	Protein (%)	0857	18.980	0.08000	18.962	0.29434	101.76	196	0.06	002.06	0
002	Protein (%)	2089	18.985	0.09000	18.962	0.29434	101.76	196	0.08	002.06	0
002	Protein (%)	0199	19.000	0.02000	18.962	0.29434	101.76	196	0.13	002.06	0
002	Protein (%)	2012	19.000	0.44000	18.962	0.29434	101.76	196	0.13	002.06	0
002	Protein (%)	0037	19.010	0.04000	18.962	0.29434	101.76	196	0.16	002.06	0
002	Protein (%)	0897	19.015	0.01000	18.962	0.29434	101.76	196	0.18	002.05	0
002	Protein (%)	0027	19.018	0.09400	18.962	0.29434	101.76	196	0.19	002.06	0
002	Protein (%)	0508	19.031	0.11300	18.962	0.29434	101.76	196	0.23	002.06	0
002	Protein (%)	2059	19.031	0.12500	18.962	0.29434	101.76	196	0.24	002.06	0
002	Protein (%)	0948	19.035	0.03000	18.962	0.29434	101.76	196	0.25	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein (%)	0553	19.040	0.16000	18.962	0.29434	101.76	196	0.27	002.06	0
002	Protein (%)	0904	19.040	0.00000	18.962	0.29434	101.76	196	0.27	002.06	0
002	Protein (%)	0884	19.050	0.30000	18.962	0.29434	101.76	196	0.30	002.06	0
002	Protein (%)	0010	19.050	0.50000	18.962	0.29434	101.76	196	0.30	002.06	0
002	Protein (%)	0016	19.050	0.10000	18.962	0.29434	101.76	196	0.30	002.06	0
002	Protein (%)	0353	19.050	0.18000	18.962	0.29434	101.76	196	0.30	002.06	0
002	Protein (%)	0843	19.050	0.18000	18.962	0.29434	101.76	196	0.30	002.06	0
002	Protein (%)	0859	19.051	0.04900	18.962	0.29434	101.76	196	0.30	002.06	0
002	Protein (%)	0089	19.060	0.00000	18.962	0.29434	101.76	196	0.33	002.06	0
002	Protein (%)	2075	19.060	0.04000	18.962	0.29434	101.76	196	0.33	002.06	0
002	Protein (%)	0233	19.075	0.01000	18.962	0.29434	101.76	196	0.38	002.06	0
002	Protein (%)	2006	19.080	0.06000	18.962	0.29434	101.76	196	0.40	002.06	0
002	Protein (%)	0893	19.085	0.01000	18.962	0.29434	101.76	196	0.42	002.05	0
002	Protein (%)	0202	19.100	0.04000	18.962	0.29434	101.76	196	0.47	002.06	0
002	Protein (%)	0354	19.100	0.02000	18.962	0.29434	101.76	196	0.47	002.06	0
002	Protein (%)	2129	19.100	0.44000	18.962	0.29434	101.76	196	0.47	002.06	0
002	Protein (%)	1019	19.101	0.16000	18.962	0.29434	101.76	196	0.47	002.06	0
002	Protein (%)	0626	19.104	0.06500	18.962	0.29434	101.76	196	0.48	002.06	0
002	Protein (%)	0035	19.105	0.13000	18.962	0.29434	101.76	196	0.49	002.06	0
002	Protein (%)	0001	19.115	0.11000	18.962	0.29434	101.76	196	0.52	002.06	0
002	Protein (%)	0190	19.120	0.18000	18.962	0.29434	101.76	196	0.54	002.06	0
002	Protein (%)	0164	19.130	0.10000	18.962	0.29434	101.76	196	0.57	002.01	0
002	Protein (%)	0142	19.140	0.02000	18.962	0.29434	101.76	196	0.61	002.06	0
002	Protein (%)	0511	19.140	0.08000	18.962	0.29434	101.76	196	0.61	002.06	0
002	Protein (%)	0300	19.149	0.01200	18.962	0.29434	101.76	196	0.64	002.06	0
002	Protein (%)	0955	19.150	0.10000	18.962	0.29434	101.76	196	0.64	002.00	0
002	Protein (%)	0152	19.150	0.10000	18.962	0.29434	101.76	196	0.64	002.01	0
002	Protein (%)	0045	19.150	0.10000	18.962	0.29434	101.76	196	0.64	002.06	0
002	Protein (%)	0413	19.150	0.10000	18.962	0.29434	101.76	196	0.64	002.06	0
002	Protein (%)	0726	19.160	0.06000	18.962	0.29434	101.76	196	0.67	002.06	0
002	Protein (%)	0563	19.170	0.00500	18.962	0.29434	101.76	196	0.71	002.06	0
002	Protein (%)	0676	19.180	0.08000	18.962	0.29434	101.76	196	0.74	002.06	0
002	Protein (%)	0512	19.190	0.28000	18.962	0.29434	101.76	196	0.78	002.06	0
002	Protein (%)	0171	19.200	0.10000	18.962	0.29434	101.76	196	0.81	002.06	0
002	Protein (%)	0964	19.202	0.00700	18.962	0.29434	101.76	196	0.81	002.06	0
002	Protein (%)	2150	19.204	0.24400	18.962	0.29434	101.76	196	0.82	002.06	0
002	Protein (%)	0098	19.210	0.16000	18.962	0.29434	101.76	196	0.84	002.06	0
002	Protein (%)	0682	19.210	0.00000	18.962	0.29434	101.76	196	0.84	002.06	0
002	Protein (%)	0644	19.220	0.06000	18.962	0.29434	101.76	196	0.88	002.99	0
002	Protein (%)	2057	19.224	0.32500	18.962	0.29434	101.76	196	0.89	002.06	0
002	Protein (%)	0541	19.225	0.23000	18.962	0.29434	101.76	196	0.89	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein (%)	0034	19.230	0.22000	18.962	0.29434	101.76	196	0.91	002.06	0
002	Protein (%)	0036	19.248	0.06480	18.962	0.29434	101.76	196	0.97	002.02	0
002	Protein (%)	0226	19.250	0.10000	18.962	0.29434	101.76	196	0.98	002.06	0
002	Protein (%)	0510	19.250	0.10000	18.962	0.29434	101.76	196	0.98	002.06	0
002	Protein (%)	0872	19.250	0.32000	18.962	0.29434	101.76	196	0.98	002.06	0
002	Protein (%)	0164	19.255	0.05000	18.962	0.29434	101.76	196	1.00	002.06	0
002	Protein (%)	0630	19.255	0.19000	18.962	0.29434	101.76	196	1.00	002.06	0
002	Protein (%)	0083	19.260	0.20000	18.962	0.29434	101.76	196	1.01	002.06	0
002	Protein (%)	0208	19.300	0.06000	18.962	0.29434	101.76	196	1.15	002.06	0
002	Protein (%)	0941	19.300	0.10000	18.962	0.29434	101.76	196	1.15	002.06	0
002	Protein (%)	0880	19.302	0.34300	18.962	0.29434	101.76	196	1.15	002.06	0
002	Protein (%)	0186	19.304	0.17100	18.962	0.29434	101.76	196	1.16	002.06	0
002	Protein (%)	0662	19.306	0.06500	18.962	0.29434	101.76	196	1.17	002.01	0
002	Protein (%)	0425	19.320	0.46000	18.962	0.29434	101.76	196	1.22	002.06	0
002	Protein (%)	0049	19.345	0.41000	18.962	0.29434	101.76	196	1.30	002.06	0
002	Protein (%)	0574	19.370	0.06000	18.962	0.29434	101.76	196	1.39	002.06	0
002	Protein (%)	0036	19.395	0.11000	18.962	0.29434	101.76	196	1.47	002.06	0
002	Protein (%)	0504	19.415	0.35000	18.962	0.29434	101.76	196	1.54	002.06	0
002	Protein (%)	2076	19.434	0.58200	18.962	0.29434	101.76	196	1.60	002.06	0
002	Protein (%)	0006	19.455	0.05000	18.962	0.29434	101.76	196	1.68	002.06	0
002	Protein (%)	0953	19.480	0.42000	18.962	0.29434	101.76	196	1.76	002.06	0
002	Protein (%)	0918	19.505	1.1900	18.962	0.29434	101.76	196	1.85	002.06	0
002	Protein (%)	0876	19.506	0.28900	18.962	0.29434	101.76	196	1.85	002.06	0
002	Protein (%)	0043	19.530	0.04000	18.962	0.29434	101.76	196	1.93	002.06	0
002	Protein (%)	0407	19.538	0.20200	18.962	0.29434	101.76	196	1.96	002.06	0
002	Protein (%)	2128	19.586	0.06100	18.962	0.29434	101.76	196	2.12	002.11	0
002	Protein (%)	0138	19.715	0.06600	18.962	0.29434	101.76	196	2.56	002.06	0
002	Protein (%)	0520	19.900	0.22000	18.962	0.29434	101.76	196	3.19	002.06	0
002	Protein (%)	0852	20.020	0.14000	18.962	0.29434	101.76	196	3.60	002.11	0
002	Protein (%)	0613	20.050	0.18000	18.962	0.29434	101.76	196	3.70	002.11	0
002	Protein (%)	0889	20.247	0.34600	18.962	0.29434	101.76	196	4.37	002.11	0
002	Protein (%)	2120	20.300	0.28000	18.962	0.29434	101.76	196	4.55	002.11	0
002	Protein (%)	2121	20.585	0.21000	18.962	0.29434	101.76	196	5.51	002.11	0
002	Protein (%)	0553	21.380	0.32000	18.962	0.29434	101.76	196	8.22	002.11	0
002	Protein (%)	0405	24.850	0.06000	18.962	0.29434	101.76	196	20.00	002.04	0
002	Protein (%)	0612	9,976.9	19,914	18.962	0.29434	101.76	196	33831.36	002.06	0
003	Fat (%)	0876	0.86500	0.59000	5.0611	0.18949	0.11779	147	-22.14	003.14	0
003	Fat (%)	0880	2.5000	1.2000	5.0611	0.18949	0.11779	147	-13.52	003.99	0
003	Fat (%)	0952	3.1200	0.02000	5.0611	0.18949	0.11779	147	-10.24	003.10	0
003	Fat (%)	0003	3.4100	0.98000	5.0611	0.18949	0.11779	147	-8.71	003.14	0
003	Fat (%)	0357	3.8000	0.00000	5.0611	0.18949	0.11779	147	-6.66	003.14	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat (%)	0940	4.3100	0.20000	5.0611	0.18949	0.11779	147	-3.96	003.12	0
003	Fat (%)	2120	4.4950	0.07000	5.0611	0.18949	0.11779	147	-2.99	003.11	0
003	Fat (%)	2152	4.6100	0.06000	5.0611	0.18949	0.11779	147	-2.38	003.12	0
003	Fat (%)	2129	4.6900	0.18000	5.0611	0.18949	0.11779	147	-1.96	003.14	0
003	Fat (%)	0034	4.7150	0.53000	5.0611	0.18949	0.11779	147	-1.83	003.14	0
003	Fat (%)	0574	4.7150	0.05000	5.0611	0.18949	0.11779	147	-1.83	003.14	0
003	Fat (%)	0298	4.7200	0.02000	5.0611	0.18949	0.11779	147	-1.80	003.10	0
003	Fat (%)	2155	4.7300	0.10000	5.0611	0.18949	0.11779	147	-1.75	003.10	0
003	Fat (%)	2022	4.7350	0.13000	5.0611	0.18949	0.11779	147	-1.72	003.10	0
003	Fat (%)	0037	4.7650	0.07000	5.0611	0.18949	0.11779	147	-1.56	003.14	0
003	Fat (%)	0619	4.8150	0.01000	5.0611	0.18949	0.11779	147	-1.30	003.10	0
003	Fat (%)	2068	4.8200	0.04000	5.0611	0.18949	0.11779	147	-1.27	003.09	0
003	Fat (%)	2073	4.8200	0.16000	5.0611	0.18949	0.11779	147	-1.27	003.10	0
003	Fat (%)	0202	4.8350	0.09000	5.0611	0.18949	0.11779	147	-1.19	003.14	0
003	Fat (%)	0581	4.8400	0.10000	5.0611	0.18949	0.11779	147	-1.17	003.14	0
003	Fat (%)	0613	4.8450	0.01000	5.0611	0.18949	0.11779	147	-1.14	003.14	0
003	Fat (%)	2121	4.8450	0.03000	5.0611	0.18949	0.11779	147	-1.14	003.11	0
003	Fat (%)	0019	4.8600	0.14000	5.0611	0.18949	0.11779	147	-1.06	003.14	0
003	Fat (%)	0144	4.8750	0.01000	5.0611	0.18949	0.11779	147	-0.98	003.14	0
003	Fat (%)	0536	4.8800	0.48000	5.0611	0.18949	0.11779	147	-0.96	003.99	0
003	Fat (%)	0861	4.8950	0.07000	5.0611	0.18949	0.11779	147	-0.88	003.10	0
003	Fat (%)	0689	4.9000	0.20000	5.0611	0.18949	0.11779	147	-0.85	003.10	0
003	Fat (%)	0208	4.9000	0.04000	5.0611	0.18949	0.11779	147	-0.85	003.13	0
003	Fat (%)	0036	4.9015	0.02840	5.0611	0.18949	0.11779	147	-0.84	003.10	0
003	Fat (%)	0615	4.9050	0.11000	5.0611	0.18949	0.11779	147	-0.82	003.00	0
003	Fat (%)	0520	4.9050	0.37000	5.0611	0.18949	0.11779	147	-0.82	003.14	0
003	Fat (%)	2075	4.9050	0.17000	5.0611	0.18949	0.11779	147	-0.82	003.14	0
003	Fat (%)	0553	4.9150	0.09000	5.0611	0.18949	0.11779	147	-0.77	003.10	0
003	Fat (%)	0265	4.9250	0.09000	5.0611	0.18949	0.11779	147	-0.72	003.14	0
003	Fat (%)	0098	4.9400	0.02000	5.0611	0.18949	0.11779	147	-0.64	003.10	0
003	Fat (%)	0242	4.9400	0.14000	5.0611	0.18949	0.11779	147	-0.64	003.10	0
003	Fat (%)	0541	4.9450	0.03000	5.0611	0.18949	0.11779	147	-0.61	003.10	0
003	Fat (%)	0358	4.9450	0.01000	5.0611	0.18949	0.11779	147	-0.61	003.14	0
003	Fat (%)	0089	4.9500	0.02000	5.0611	0.18949	0.11779	147	-0.59	003.14	0
003	Fat (%)	0183	4.9620	0.01000	5.0611	0.18949	0.11779	147	-0.52	003.14	0
003	Fat (%)	0142	4.9650	0.05000	5.0611	0.18949	0.11779	147	-0.51	003.00	0
003	Fat (%)	2012	4.9650	0.01000	5.0611	0.18949	0.11779	147	-0.51	003.09	0
003	Fat (%)	0598	4.9700	0.00000	5.0611	0.18949	0.11779	147	-0.48	003.14	0
003	Fat (%)	0939	4.9700	0.06000	5.0611	0.18949	0.11779	147	-0.48	003.00	0
003	Fat (%)	0693	4.9800	0.00000	5.0611	0.18949	0.11779	147	-0.43	003.10	0
003	Fat (%)	0407	4.9824	0.17450	5.0611	0.18949	0.11779	147	-0.42	003.14	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat (%)	0613	4.9900	0.02000	5.0611	0.18949	0.11779	147	-0.38	003.11	0
003	Fat (%)	0843	4.9900	0.12000	5.0611	0.18949	0.11779	147	-0.38	003.14	0
003	Fat (%)	0425	4.9950	0.05000	5.0611	0.18949	0.11779	147	-0.35	003.06	0
003	Fat (%)	0226	5.0000	0.00000	5.0611	0.18949	0.11779	147	-0.32	003.09	0
003	Fat (%)	0956	5.0000	0.00000	5.0611	0.18949	0.11779	147	-0.32	003.14	0
003	Fat (%)	0662	5.0048	0.52390	5.0611	0.18949	0.11779	147	-0.30	003.10	0
003	Fat (%)	0148	5.0050	0.07000	5.0611	0.18949	0.11779	147	-0.30	003.06	0
003	Fat (%)	0199	5.0100	0.04000	5.0611	0.18949	0.11779	147	-0.27	003.06	0
003	Fat (%)	0263	5.0115	0.01700	5.0611	0.18949	0.11779	147	-0.26	003.09	0
003	Fat (%)	0674	5.0200	0.00000	5.0611	0.18949	0.11779	147	-0.22	003.09	0
003	Fat (%)	0027	5.0215	0.03300	5.0611	0.18949	0.11779	147	-0.21	003.09	0
003	Fat (%)	0004	5.0250	0.01000	5.0611	0.18949	0.11779	147	-0.19	003.09	0
003	Fat (%)	0169	5.0350	0.05000	5.0611	0.18949	0.11779	147	-0.14	003.06	0
003	Fat (%)	2109	5.0400	0.06000	5.0611	0.18949	0.11779	147	-0.11	003.00	0
003	Fat (%)	0897	5.0400	0.02000	5.0611	0.18949	0.11779	147	-0.11	003.10	0
003	Fat (%)	0294	5.0450	0.01000	5.0611	0.18949	0.11779	147	-0.09	003.06	0
003	Fat (%)	0948	5.0450	0.01000	5.0611	0.18949	0.11779	147	-0.09	003.09	0
003	Fat (%)	0045	5.0450	0.05000	5.0611	0.18949	0.11779	147	-0.09	003.10	0
003	Fat (%)	0893	5.0450	0.01000	5.0611	0.18949	0.11779	147	-0.09	003.12	0
003	Fat (%)	0187	5.0450	0.05000	5.0611	0.18949	0.11779	147	-0.09	003.13	0
003	Fat (%)	0004	5.0450	0.03000	5.0611	0.18949	0.11779	147	-0.09	003.14	0
003	Fat (%)	0108	5.0450	0.05000	5.0611	0.18949	0.11779	147	-0.09	003.14	0
003	Fat (%)	0098	5.0550	0.05000	5.0611	0.18949	0.11779	147	-0.03	003.09	0
003	Fat (%)	0903	5.0550	0.01000	5.0611	0.18949	0.11779	147	-0.03	003.12	0
003	Fat (%)	0175	5.0550	0.23000	5.0611	0.18949	0.11779	147	-0.03	003.14	0
003	Fat (%)	0529	5.0550	0.03000	5.0611	0.18949	0.11779	147	-0.03	003.14	0
003	Fat (%)	0968	5.0560	0.10200	5.0611	0.18949	0.11779	147	-0.03	003.14	0
003	Fat (%)	0563	5.0599	0.03620	5.0611	0.18949	0.11779	147	-0.01	003.06	0
003	Fat (%)	0629	5.0600	0.12000	5.0611	0.18949	0.11779	147	-0.01	003.10	0
003	Fat (%)	0098	5.0600	0.22000	5.0611	0.18949	0.11779	147	-0.01	003.13	0
003	Fat (%)	0066	5.0650	0.11000	5.0611	0.18949	0.11779	147	0.02	003.14	0
003	Fat (%)	0553	5.0700	0.02000	5.0611	0.18949	0.11779	147	0.05	003.11	0
003	Fat (%)	0353	5.0750	0.11000	5.0611	0.18949	0.11779	147	0.07	003.09	0
003	Fat (%)	0100	5.0750	0.15000	5.0611	0.18949	0.11779	147	0.07	003.10	0
003	Fat (%)	0852	5.0900	0.02000	5.0611	0.18949	0.11779	147	0.15	003.11	0
003	Fat (%)	0033	5.0900	0.30000	5.0611	0.18949	0.11779	147	0.15	003.13	0
003	Fat (%)	0918	5.0950	0.03000	5.0611	0.18949	0.11779	147	0.18	003.06	0
003	Fat (%)	0345	5.1000	0.10000	5.0611	0.18949	0.11779	147	0.21	003.00	0
003	Fat (%)	0233	5.1000	0.00000	5.0611	0.18949	0.11779	147	0.21	003.10	0
003	Fat (%)	0366	5.1000	0.20000	5.0611	0.18949	0.11779	147	0.21	003.10	0
003	Fat (%)	0009	5.1000	0.06000	5.0611	0.18949	0.11779	147	0.21	003.14	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat (%)	0653	5.1050	0.01000	5.0611	0.18949	0.11779	147	0.23	003.00	0
003	Fat (%)	0878	5.1050	0.01000	5.0611	0.18949	0.11779	147	0.23	003.10	0
003	Fat (%)	0644	5.1050	0.41000	5.0611	0.18949	0.11779	147	0.23	003.99	0
003	Fat (%)	0656	5.1100	0.04000	5.0611	0.18949	0.11779	147	0.26	003.09	0
003	Fat (%)	0646	5.1100	0.08000	5.0611	0.18949	0.11779	147	0.26	003.13	0
003	Fat (%)	2009	5.1206	0.01510	5.0611	0.18949	0.11779	147	0.31	003.10	0
003	Fat (%)	0205	5.1240	0.04200	5.0611	0.18949	0.11779	147	0.33	003.13	0
003	Fat (%)	0895	5.1250	0.01000	5.0611	0.18949	0.11779	147	0.34	003.12	0
003	Fat (%)	0001	5.1255	0.27500	5.0611	0.18949	0.11779	147	0.34	003.14	0
003	Fat (%)	0682	5.1400	0.00000	5.0611	0.18949	0.11779	147	0.42	003.06	0
003	Fat (%)	0910	5.1400	0.00000	5.0611	0.18949	0.11779	147	0.42	003.06	0
003	Fat (%)	0032	5.1500	0.10000	5.0611	0.18949	0.11779	147	0.47	003.00	0
003	Fat (%)	0689	5.1500	0.10000	5.0611	0.18949	0.11779	147	0.47	003.06	0
003	Fat (%)	0685	5.1500	0.12000	5.0611	0.18949	0.11779	147	0.47	003.09	0
003	Fat (%)	0015	5.1500	0.04000	5.0611	0.18949	0.11779	147	0.47	003.14	0
003	Fat (%)	0229	5.1500	0.02000	5.0611	0.18949	0.11779	147	0.47	003.14	0
003	Fat (%)	0278	5.1500	0.10000	5.0611	0.18949	0.11779	147	0.47	003.14	0
003	Fat (%)	0618	5.1520	0.15400	5.0611	0.18949	0.11779	147	0.48	003.10	0
003	Fat (%)	0229	5.1550	0.11000	5.0611	0.18949	0.11779	147	0.50	003.06	0
003	Fat (%)	0849	5.1590	0.08200	5.0611	0.18949	0.11779	147	0.52	003.06	0
003	Fat (%)	0511	5.1600	0.04000	5.0611	0.18949	0.11779	147	0.52	003.06	0
003	Fat (%)	0505	5.1600	0.00000	5.0611	0.18949	0.11779	147	0.52	003.09	0
003	Fat (%)	0049	5.1600	0.20000	5.0611	0.18949	0.11779	147	0.52	003.14	0
003	Fat (%)	0354	5.1650	0.01000	5.0611	0.18949	0.11779	147	0.55	003.09	0
003	Fat (%)	0630	5.1650	0.01000	5.0611	0.18949	0.11779	147	0.55	003.09	0
003	Fat (%)	0300	5.1745	0.03300	5.0611	0.18949	0.11779	147	0.60	003.06	0
003	Fat (%)	0083	5.1750	0.09000	5.0611	0.18949	0.11779	147	0.60	003.06	0
003	Fat (%)	2062	5.1764	0.01330	5.0611	0.18949	0.11779	147	0.61	003.12	0
003	Fat (%)	0619	5.1850	0.01000	5.0611	0.18949	0.11779	147	0.65	003.14	0
003	Fat (%)	0692	5.2000	0.00000	5.0611	0.18949	0.11779	147	0.73	003.14	0
003	Fat (%)	0508	5.2010	0.13260	5.0611	0.18949	0.11779	147	0.74	003.09	0
003	Fat (%)	0552	5.2050	0.15000	5.0611	0.18949	0.11779	147	0.76	003.06	0
003	Fat (%)	0016	5.2050	0.01000	5.0611	0.18949	0.11779	147	0.76	003.99	0
003	Fat (%)	2006	5.2100	0.14000	5.0611	0.18949	0.11779	147	0.79	003.13	0
003	Fat (%)	2128	5.2127	0.08120	5.0611	0.18949	0.11779	147	0.80	003.11	0
003	Fat (%)	0885	5.2395	0.32400	5.0611	0.18949	0.11779	147	0.94	003.00	0
003	Fat (%)	0175	5.2400	0.08000	5.0611	0.18949	0.11779	147	0.94	003.00	0
003	Fat (%)	0955	5.2400	0.08000	5.0611	0.18949	0.11779	147	0.94	003.14	0
003	Fat (%)	0889	5.2750	0.03000	5.0611	0.18949	0.11779	147	1.13	003.10	0
003	Fat (%)	0722	5.2802	0.03820	5.0611	0.18949	0.11779	147	1.16	003.09	0
003	Fat (%)	0948	5.2900	0.18000	5.0611	0.18949	0.11779	147	1.21	003.14	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat (%)	0510	5.3000	0.00000	5.0611	0.18949	0.11779	147	1.26	003.09	0
003	Fat (%)	0171	5.3050	0.11000	5.0611	0.18949	0.11779	147	1.29	003.12	0
003	Fat (%)	0190	5.3050	0.07000	5.0611	0.18949	0.11779	147	1.29	003.14	0
003	Fat (%)	2053	5.3100	0.04000	5.0611	0.18949	0.11779	147	1.31	003.00	0
003	Fat (%)	0038	5.3150	0.09000	5.0611	0.18949	0.11779	147	1.34	003.09	0
003	Fat (%)	0675	5.3200	0.50000	5.0611	0.18949	0.11779	147	1.37	003.09	0
003	Fat (%)	0726	5.3200	0.00000	5.0611	0.18949	0.11779	147	1.37	003.14	0
003	Fat (%)	0884	5.3500	0.10000	5.0611	0.18949	0.11779	147	1.52	003.06	0
003	Fat (%)	0413	5.4000	0.40000	5.0611	0.18949	0.11779	147	1.79	003.14	0
003	Fat (%)	2089	5.4050	0.31000	5.0611	0.18949	0.11779	147	1.81	003.13	0
003	Fat (%)	0504	5.4500	0.80000	5.0611	0.18949	0.11779	147	2.05	003.01	0
003	Fat (%)	0152	5.4500	0.10000	5.0611	0.18949	0.11779	147	2.05	003.09	0
003	Fat (%)	0590	5.4500	0.00000	5.0611	0.18949	0.11779	147	2.05	003.09	0
003	Fat (%)	0889	5.4600	0.14000	5.0611	0.18949	0.11779	147	2.10	003.11	0
003	Fat (%)	0676	5.5000	0.02000	5.0611	0.18949	0.11779	147	2.32	003.10	0
003	Fat (%)	0941	5.5300	0.20000	5.0611	0.18949	0.11779	147	2.47	003.99	0
003	Fat (%)	0848	5.6650	0.25000	5.0611	0.18949	0.11779	147	3.19	003.14	0
003	Fat (%)	0852	5.9400	0.12000	5.0611	0.18949	0.11779	147	4.64	003.06	0
003	Fat (%)	1019	7.7250	0.61000	5.0611	0.18949	0.11779	147	14.06	003.99	0
004	Fiber (%)	0265	3.5250	0.09000	4.7798	0.52002	0.22130	126	-2.41	004.07	0
004	Fiber (%)	0708	3.7850	0.11000	4.7798	0.52002	0.22130	126	-1.91	004.07	0
004	Fiber (%)	0910	4.0000	0.00000	4.7798	0.52002	0.22130	126	-1.50	004.07	0
004	Fiber (%)	0035	4.0450	0.23000	4.7798	0.52002	0.22130	126	-1.41	004.07	0
004	Fiber (%)	0619	4.0480	0.24400	4.7798	0.52002	0.22130	126	-1.41	004.07	0
004	Fiber (%)	0895	4.0500	0.02000	4.7798	0.52002	0.22130	126	-1.40	004.06	0
004	Fiber (%)	0026	4.0500	0.10000	4.7798	0.52002	0.22130	126	-1.40	004.07	0
004	Fiber (%)	0921	4.0500	0.04000	4.7798	0.52002	0.22130	126	-1.40	004.03	0
004	Fiber (%)	0644	4.0900	0.18000	4.7798	0.52002	0.22130	126	-1.33	004.99	0
004	Fiber (%)	0629	4.1000	0.00000	4.7798	0.52002	0.22130	126	-1.31	004.99	0
004	Fiber (%)	0726	4.1600	0.12000	4.7798	0.52002	0.22130	126	-1.19	004.07	0
004	Fiber (%)	0952	4.1750	0.01000	4.7798	0.52002	0.22130	126	-1.16	004.06	0
004	Fiber (%)	0889	4.1850	0.23000	4.7798	0.52002	0.22130	126	-1.14	004.11	0
004	Fiber (%)	0653	4.2250	0.03000	4.7798	0.52002	0.22130	126	-1.07	004.07	0
004	Fiber (%)	0169	4.2550	0.03000	4.7798	0.52002	0.22130	126	-1.01	004.00	0
004	Fiber (%)	0848	4.2650	0.69000	4.7798	0.52002	0.22130	126	-0.99	004.07	0
004	Fiber (%)	0009	4.2750	0.15000	4.7798	0.52002	0.22130	126	-0.97	004.07	0
004	Fiber (%)	0003	4.2800	0.58000	4.7798	0.52002	0.22130	126	-0.96	004.07	0
004	Fiber (%)	0413	4.3000	0.20000	4.7798	0.52002	0.22130	126	-0.92	004.07	0
004	Fiber (%)	0098	4.3050	0.05000	4.7798	0.52002	0.22130	126	-0.91	004.06	0
004	Fiber (%)	0298	4.3100	0.02000	4.7798	0.52002	0.22130	126	-0.90	004.00	0
004	Fiber (%)	0590	4.3300	0.06000	4.7798	0.52002	0.22130	126	-0.86	004.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber (%)	2053	4.3450	0.01000	4.7798	0.52002	0.22130	126	-0.84	004.07	0
004	Fiber (%)	0884	4.3500	0.10000	4.7798	0.52002	0.22130	126	-0.83	004.00	0
004	Fiber (%)	0610	4.3500	0.10000	4.7798	0.52002	0.22130	126	-0.83	004.99	0
004	Fiber (%)	0505	4.3600	0.02000	4.7798	0.52002	0.22130	126	-0.81	004.07	0
004	Fiber (%)	0066	4.3700	0.20000	4.7798	0.52002	0.22130	126	-0.79	004.07	0
004	Fiber (%)	0903	4.3850	0.01000	4.7798	0.52002	0.22130	126	-0.76	004.06	0
004	Fiber (%)	0345	4.3950	0.09000	4.7798	0.52002	0.22130	126	-0.74	004.00	0
004	Fiber (%)	0098	4.4050	0.01000	4.7798	0.52002	0.22130	126	-0.72	004.07	0
004	Fiber (%)	0940	4.4100	0.00000	4.7798	0.52002	0.22130	126	-0.71	004.99	0
004	Fiber (%)	0100	4.4250	0.01000	4.7798	0.52002	0.22130	126	-0.68	004.07	0
004	Fiber (%)	2121	4.4350	0.13000	4.7798	0.52002	0.22130	126	-0.66	004.11	0
004	Fiber (%)	0852	4.4450	0.07000	4.7798	0.52002	0.22130	126	-0.64	004.11	0
004	Fiber (%)	2089	4.4500	2.2200	4.7798	0.52002	0.22130	126	-0.63	004.03	0
004	Fiber (%)	2022	4.4600	0.04000	4.7798	0.52002	0.22130	126	-0.61	004.06	0
004	Fiber (%)	0175	4.4750	0.05000	4.7798	0.52002	0.22130	126	-0.59	004.00	0
004	Fiber (%)	0885	4.4795	0.06750	4.7798	0.52002	0.22130	126	-0.58	004.06	0
004	Fiber (%)	2062	4.4880	0.09800	4.7798	0.52002	0.22130	126	-0.56	004.06	0
004	Fiber (%)	0032	4.5000	0.00000	4.7798	0.52002	0.22130	126	-0.54	004.07	0
004	Fiber (%)	0294	4.5000	0.00000	4.7798	0.52002	0.22130	126	-0.54	004.07	0
004	Fiber (%)	0956	4.5000	0.00000	4.7798	0.52002	0.22130	126	-0.54	004.07	0
004	Fiber (%)	0553	4.5050	0.53000	4.7798	0.52002	0.22130	126	-0.53	004.07	0
004	Fiber (%)	2073	4.5050	0.13000	4.7798	0.52002	0.22130	126	-0.53	004.00	0
004	Fiber (%)	0897	4.5350	0.01000	4.7798	0.52002	0.22130	126	-0.47	004.06	0
004	Fiber (%)	0164	4.5500	0.10000	4.7798	0.52002	0.22130	126	-0.44	004.00	0
004	Fiber (%)	0425	4.5500	0.50000	4.7798	0.52002	0.22130	126	-0.44	004.00	0
004	Fiber (%)	0353	4.5550	0.65000	4.7798	0.52002	0.22130	126	-0.43	004.03	0
004	Fiber (%)	2076	4.5650	1.0300	4.7798	0.52002	0.22130	126	-0.41	004.00	0
004	Fiber (%)	0675	4.5650	0.11000	4.7798	0.52002	0.22130	126	-0.41	004.06	0
004	Fiber (%)	0941	4.5700	0.24000	4.7798	0.52002	0.22130	126	-0.40	004.06	0
004	Fiber (%)	0878	4.5750	0.05000	4.7798	0.52002	0.22130	126	-0.39	004.06	0
004	Fiber (%)	0074	4.5750	0.01000	4.7798	0.52002	0.22130	126	-0.39	004.07	0
004	Fiber (%)	0682	4.6000	0.00000	4.7798	0.52002	0.22130	126	-0.35	004.07	0
004	Fiber (%)	0552	4.6050	0.01000	4.7798	0.52002	0.22130	126	-0.34	004.07	0
004	Fiber (%)	0033	4.6300	0.04000	4.7798	0.52002	0.22130	126	-0.29	004.07	0
004	Fiber (%)	0948	4.6550	0.03000	4.7798	0.52002	0.22130	126	-0.24	004.06	0
004	Fiber (%)	0520	4.6550	0.35000	4.7798	0.52002	0.22130	126	-0.24	004.07	0
004	Fiber (%)	2097	4.6590	0.09600	4.7798	0.52002	0.22130	126	-0.23	004.06	0
004	Fiber (%)	2120	4.6850	0.17000	4.7798	0.52002	0.22130	126	-0.18	004.11	0
004	Fiber (%)	0171	4.6950	0.07000	4.7798	0.52002	0.22130	126	-0.16	004.00	0
004	Fiber (%)	0208	4.7150	0.07000	4.7798	0.52002	0.22130	126	-0.12	004.00	0
004	Fiber (%)	0689	4.7500	0.30000	4.7798	0.52002	0.22130	126	-0.06	004.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber (%)	0205	4.7600	0.08000	4.7798	0.52002	0.22130	126	-0.04	004.06	0
004	Fiber (%)	0613	4.7800	0.06000	4.7798	0.52002	0.22130	126	0.00	004.11	0
004	Fiber (%)	0918	4.7850	0.55000	4.7798	0.52002	0.22130	126	0.01	004.06	0
004	Fiber (%)	0354	4.7900	0.10000	4.7798	0.52002	0.22130	126	0.02	004.07	0
004	Fiber (%)	0529	4.7950	0.33000	4.7798	0.52002	0.22130	126	0.03	004.07	0
004	Fiber (%)	0646	4.7950	0.15000	4.7798	0.52002	0.22130	126	0.03	004.07	0
004	Fiber (%)	0027	4.8005	0.04100	4.7798	0.52002	0.22130	126	0.04	004.06	0
004	Fiber (%)	2109	4.8400	0.10000	4.7798	0.52002	0.22130	126	0.12	004.07	0
004	Fiber (%)	0511	4.8500	0.10000	4.7798	0.52002	0.22130	126	0.14	004.00	0
004	Fiber (%)	0722	4.8511	0.03150	4.7798	0.52002	0.22130	126	0.14	004.06	0
004	Fiber (%)	2012	4.8700	0.10000	4.7798	0.52002	0.22130	126	0.17	004.07	0
004	Fiber (%)	0034	4.8800	0.02000	4.7798	0.52002	0.22130	126	0.19	004.07	0
004	Fiber (%)	0948	4.8800	0.10000	4.7798	0.52002	0.22130	126	0.19	004.07	0
004	Fiber (%)	0199	4.9000	0.10000	4.7798	0.52002	0.22130	126	0.23	004.06	0
004	Fiber (%)	0190	4.9000	0.20000	4.7798	0.52002	0.22130	126	0.23	004.07	0
004	Fiber (%)	2075	4.9000	0.60000	4.7798	0.52002	0.22130	126	0.23	004.07	0
004	Fiber (%)	0553	4.9050	0.07000	4.7798	0.52002	0.22130	126	0.24	004.11	0
004	Fiber (%)	0510	4.9500	0.30000	4.7798	0.52002	0.22130	126	0.33	004.07	0
004	Fiber (%)	0226	4.9500	0.10000	4.7798	0.52002	0.22130	126	0.33	004.00	0
004	Fiber (%)	0674	4.9500	0.00000	4.7798	0.52002	0.22130	126	0.33	004.06	0
004	Fiber (%)	2152	4.9650	0.03000	4.7798	0.52002	0.22130	126	0.36	004.06	0
004	Fiber (%)	0183	4.9800	0.02000	4.7798	0.52002	0.22130	126	0.38	004.00	0
004	Fiber (%)	0017	5.0000	0.20000	4.7798	0.52002	0.22130	126	0.42	004.00	0
004	Fiber (%)	0904	5.0050	0.05000	4.7798	0.52002	0.22130	126	0.43	004.07	0
004	Fiber (%)	0004	5.0200	0.10000	4.7798	0.52002	0.22130	126	0.46	004.07	0
004	Fiber (%)	0083	5.0400	0.30000	4.7798	0.52002	0.22130	126	0.50	004.07	0
004	Fiber (%)	0242	5.0550	0.03000	4.7798	0.52002	0.22130	126	0.53	004.07	0
004	Fiber (%)	2149	5.0700	0.46000	4.7798	0.52002	0.22130	126	0.56	004.06	0
004	Fiber (%)	0202	5.1050	0.07000	4.7798	0.52002	0.22130	126	0.63	004.07	0
004	Fiber (%)	0038	5.1150	0.07000	4.7798	0.52002	0.22130	126	0.64	004.07	0
004	Fiber (%)	0263	5.1215	0.05700	4.7798	0.52002	0.22130	126	0.66	004.07	0
004	Fiber (%)	0536	5.1300	0.00000	4.7798	0.52002	0.22130	126	0.67	004.07	0
004	Fiber (%)	0354	5.1450	0.07000	4.7798	0.52002	0.22130	126	0.70	004.00	0
004	Fiber (%)	0229	5.1500	0.30000	4.7798	0.52002	0.22130	126	0.71	004.07	0
004	Fiber (%)	0939	5.2000	0.30000	4.7798	0.52002	0.22130	126	0.81	004.07	0
004	Fiber (%)	2006	5.2150	0.19000	4.7798	0.52002	0.22130	126	0.84	004.06	0
004	Fiber (%)	0028	5.2500	0.50000	4.7798	0.52002	0.22130	126	0.90	004.07	0
004	Fiber (%)	0581	5.2500	0.10000	4.7798	0.52002	0.22130	126	0.90	004.07	0
004	Fiber (%)	0692	5.2500	0.70000	4.7798	0.52002	0.22130	126	0.90	004.07	0
004	Fiber (%)	0968	5.2555	0.00900	4.7798	0.52002	0.22130	126	0.91	004.06	0
004	Fiber (%)	0968	5.2555	0.21100	4.7798	0.52002	0.22130	126	0.91	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber (%)	0849	5.2850	0.01000	4.7798	0.52002	0.22130	126	0.97	004.07	0
004	Fiber (%)	0626	5.3400	0.20000	4.7798	0.52002	0.22130	126	1.08	004.03	0
004	Fiber (%)	0643	5.3550	0.33000	4.7798	0.52002	0.22130	126	1.11	004.07	0
004	Fiber (%)	0563	5.3602	0.39550	4.7798	0.52002	0.22130	126	1.12	004.07	0
004	Fiber (%)	0613	5.3650	0.61000	4.7798	0.52002	0.22130	126	1.13	004.07	0
004	Fiber (%)	0861	5.3750	0.67000	4.7798	0.52002	0.22130	126	1.14	004.07	0
004	Fiber (%)	2128	5.3883	0.01910	4.7798	0.52002	0.22130	126	1.17	004.11	0
004	Fiber (%)	0676	5.5150	0.01000	4.7798	0.52002	0.22130	126	1.41	004.06	0
004	Fiber (%)	0893	5.5900	0.02000	4.7798	0.52002	0.22130	126	1.56	004.06	0
004	Fiber (%)	0045	5.6250	0.01000	4.7798	0.52002	0.22130	126	1.63	004.03	0
004	Fiber (%)	0598	5.6850	0.57000	4.7798	0.52002	0.22130	126	1.74	004.07	0
004	Fiber (%)	0089	5.7900	0.02000	4.7798	0.52002	0.22130	126	1.94	004.07	0
004	Fiber (%)	0016	5.9400	0.00000	4.7798	0.52002	0.22130	126	2.23	004.07	0
004	Fiber (%)	0693	5.9650	0.31000	4.7798	0.52002	0.22130	126	2.28	004.03	0
004	Fiber (%)	0366	6.0500	0.30000	4.7798	0.52002	0.22130	126	2.44	004.01	0
004	Fiber (%)	0278	6.1000	0.40000	4.7798	0.52002	0.22130	126	2.54	004.07	0
004	Fiber (%)	0876	6.1700	1.4400	4.7798	0.52002	0.22130	126	2.67	004.07	0
004	Fiber (%)	0019	6.5750	0.75000	4.7798	0.52002	0.22130	126	3.45	004.07	0
004	Fiber (%)	0407	7.6640	0.05580	4.7798	0.52002	0.22130	126	5.55	004.07	0
004	Fiber (%)	0504	9.3050	3.2100	4.7798	0.52002	0.22130	126	8.70	004.00	0
004	Fiber (%)	2009	10.065	0.13870	4.7798	0.52002	0.22130	126	10.16	004.07	0
004	Fiber (%)	0144	15.695	0.61000	4.7798	0.52002	0.22130	126	20.99	004.07	0
005	Ash (%)	0889	6.9800	0.50000	11.863	0.37506	0.15506	141	-13.02	005.11	0
005	Ash (%)	0169	8.8250	0.21000	11.863	0.37506	0.15506	141	-8.10	005.00	0
005	Ash (%)	0278	9.0600	0.30000	11.863	0.37506	0.15506	141	-7.47	005.00	0
005	Ash (%)	0425	9.0750	0.09000	11.863	0.37506	0.15506	141	-7.43	005.00	0
005	Ash (%)	0539	9.1400	0.12000	11.863	0.37506	0.15506	141	-7.26	005.00	0
005	Ash (%)	0552	9.4300	0.24000	11.863	0.37506	0.15506	141	-6.49	005.00	0
005	Ash (%)	2128	9.4392	0.54590	11.863	0.37506	0.15506	141	-6.46	005.11	0
005	Ash (%)	0613	9.5250	0.25000	11.863	0.37506	0.15506	141	-6.23	005.11	0
005	Ash (%)	0662	9.8695	0.20630	11.863	0.37506	0.15506	141	-5.32	005.99	0
005	Ash (%)	0019	10.180	2.5600	11.863	0.37506	0.15506	141	-4.49	005.00	0
005	Ash (%)	0511	10.440	0.18000	11.863	0.37506	0.15506	141	-3.79	005.00	0
005	Ash (%)	2073	10.470	0.18000	11.863	0.37506	0.15506	141	-3.71	005.00	0
005	Ash (%)	0650	10.665	0.01000	11.863	0.37506	0.15506	141	-3.19	005.00	0
005	Ash (%)	0208	10.700	0.20000	11.863	0.37506	0.15506	141	-3.10	005.00	0
005	Ash (%)	0598	10.765	0.11000	11.863	0.37506	0.15506	141	-2.93	005.00	0
005	Ash (%)	0003	10.900	0.40000	11.863	0.37506	0.15506	141	-2.57	005.05	0
005	Ash (%)	0852	10.940	0.18000	11.863	0.37506	0.15506	141	-2.46	005.11	0
005	Ash (%)	2053	11.160	0.02000	11.863	0.37506	0.15506	141	-1.88	005.00	0
005	Ash (%)	0038	11.265	0.65000	11.863	0.37506	0.15506	141	-1.60	005.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0049	11.300	1.1800	11.863	0.37506	0.15506	141	-1.50	005.00	0
005	Ash (%)	0541	11.305	0.03000	11.863	0.37506	0.15506	141	-1.49	005.00	0
005	Ash (%)	0089	11.355	0.01000	11.863	0.37506	0.15506	141	-1.36	005.00	0
005	Ash (%)	0205	11.375	0.07000	11.863	0.37506	0.15506	141	-1.30	005.00	0
005	Ash (%)	0148	11.378	0.06560	11.863	0.37506	0.15506	141	-1.29	005.00	0
005	Ash (%)	0520	11.400	0.18000	11.863	0.37506	0.15506	141	-1.24	005.00	0
005	Ash (%)	0004	11.480	0.08000	11.863	0.37506	0.15506	141	-1.02	005.00	0
005	Ash (%)	0918	11.535	0.35000	11.863	0.37506	0.15506	141	-0.88	005.00	0
005	Ash (%)	0574	11.545	0.19000	11.863	0.37506	0.15506	141	-0.85	005.99	0
005	Ash (%)	0848	11.550	0.14000	11.863	0.37506	0.15506	141	-0.84	005.00	0
005	Ash (%)	0956	11.550	0.10000	11.863	0.37506	0.15506	141	-0.84	005.00	0
005	Ash (%)	0413	11.550	0.70000	11.863	0.37506	0.15506	141	-0.84	005.05	0
005	Ash (%)	0026	11.560	0.22000	11.863	0.37506	0.15506	141	-0.81	005.00	0
005	Ash (%)	0682	11.560	0.00000	11.863	0.37506	0.15506	141	-0.81	005.00	0
005	Ash (%)	2097	11.627	0.21790	11.863	0.37506	0.15506	141	-0.63	005.00	0
005	Ash (%)	0693	11.650	0.10000	11.863	0.37506	0.15506	141	-0.57	005.00	0
005	Ash (%)	0726	11.655	0.05000	11.863	0.37506	0.15506	141	-0.56	005.05	0
005	Ash (%)	0015	11.670	0.02000	11.863	0.37506	0.15506	141	-0.52	005.00	0
005	Ash (%)	1019	11.675	0.13000	11.863	0.37506	0.15506	141	-0.50	005.00	0
005	Ash (%)	0536	11.680	0.04000	11.863	0.37506	0.15506	141	-0.49	005.99	0
005	Ash (%)	0144	11.685	0.09000	11.863	0.37506	0.15506	141	-0.48	005.00	0
005	Ash (%)	0358	11.705	0.21000	11.863	0.37506	0.15506	141	-0.42	005.00	0
005	Ash (%)	0953	11.710	0.20000	11.863	0.37506	0.15506	141	-0.41	005.00	0
005	Ash (%)	0039	11.726	0.00660	11.863	0.37506	0.15506	141	-0.37	005.00	0
005	Ash (%)	0504	11.735	0.79000	11.863	0.37506	0.15506	141	-0.34	005.00	0
005	Ash (%)	0100	11.750	0.12000	11.863	0.37506	0.15506	141	-0.30	005.00	0
005	Ash (%)	0620	11.758	0.03690	11.863	0.37506	0.15506	141	-0.28	005.00	0
005	Ash (%)	0001	11.765	0.03920	11.863	0.37506	0.15506	141	-0.26	005.00	0
005	Ash (%)	0510	11.770	0.04000	11.863	0.37506	0.15506	141	-0.25	005.00	0
005	Ash (%)	0298	11.780	0.02000	11.863	0.37506	0.15506	141	-0.22	005.00	0
005	Ash (%)	0529	11.780	0.74000	11.863	0.37506	0.15506	141	-0.22	005.00	0
005	Ash (%)	2012	11.785	0.03000	11.863	0.37506	0.15506	141	-0.21	005.00	0
005	Ash (%)	0108	11.795	0.01000	11.863	0.37506	0.15506	141	-0.18	005.00	0
005	Ash (%)	0033	11.800	0.00000	11.863	0.37506	0.15506	141	-0.17	005.00	0
005	Ash (%)	0142	11.815	0.11000	11.863	0.37506	0.15506	141	-0.13	005.00	0
005	Ash (%)	0618	11.851	0.00140	11.863	0.37506	0.15506	141	-0.03	005.00	0
005	Ash (%)	2022	11.855	0.15000	11.863	0.37506	0.15506	141	-0.02	005.00	0
005	Ash (%)	0553	11.870	0.02000	11.863	0.37506	0.15506	141	0.02	005.00	0
005	Ash (%)	0229	11.875	0.15000	11.863	0.37506	0.15506	141	0.03	005.00	0
005	Ash (%)	0035	11.885	0.03000	11.863	0.37506	0.15506	141	0.06	005.00	0
005	Ash (%)	0940	11.885	0.03000	11.863	0.37506	0.15506	141	0.06	005.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0045	11.900	0.20000	11.863	0.37506	0.15506	141	0.10	005.00	0
005	Ash (%)	0083	11.900	0.04000	11.863	0.37506	0.15506	141	0.10	005.00	0
005	Ash (%)	0674	11.900	0.00000	11.863	0.37506	0.15506	141	0.10	005.00	0
005	Ash (%)	0904	11.900	0.14000	11.863	0.37506	0.15506	141	0.10	005.05	0
005	Ash (%)	2062	11.904	0.17030	11.863	0.37506	0.15506	141	0.11	005.05	0
005	Ash (%)	0164	11.905	0.01000	11.863	0.37506	0.15506	141	0.11	005.00	0
005	Ash (%)	0505	11.910	0.06000	11.863	0.37506	0.15506	141	0.12	005.00	0
005	Ash (%)	0630	11.910	0.06000	11.863	0.37506	0.15506	141	0.12	005.00	0
005	Ash (%)	0199	11.920	0.10000	11.863	0.37506	0.15506	141	0.15	005.00	0
005	Ash (%)	0874	11.920	0.34000	11.863	0.37506	0.15506	141	0.15	005.00	0
005	Ash (%)	0263	11.925	0.01000	11.863	0.37506	0.15506	141	0.16	005.05	0
005	Ash (%)	0354	11.940	0.04000	11.863	0.37506	0.15506	141	0.20	005.00	0
005	Ash (%)	0878	11.945	0.01000	11.863	0.37506	0.15506	141	0.22	005.05	0
005	Ash (%)	0017	11.950	0.10000	11.863	0.37506	0.15506	141	0.23	005.00	0
005	Ash (%)	0226	11.950	0.30000	11.863	0.37506	0.15506	141	0.23	005.00	0
005	Ash (%)	0619	11.950	0.10000	11.863	0.37506	0.15506	141	0.23	005.05	0
005	Ash (%)	0186	11.957	0.04740	11.863	0.37506	0.15506	141	0.25	005.00	0
005	Ash (%)	0589	11.960	0.16000	11.863	0.37506	0.15506	141	0.26	005.00	0
005	Ash (%)	0893	11.960	0.00000	11.863	0.37506	0.15506	141	0.26	005.05	0
005	Ash (%)	0407	11.967	0.06990	11.863	0.37506	0.15506	141	0.28	005.00	0
005	Ash (%)	0098	11.975	0.07000	11.863	0.37506	0.15506	141	0.30	005.00	0
005	Ash (%)	0968	11.982	0.04100	11.863	0.37506	0.15506	141	0.32	005.00	0
005	Ash (%)	0653	11.985	0.07000	11.863	0.37506	0.15506	141	0.32	005.00	0
005	Ash (%)	0897	11.985	0.01000	11.863	0.37506	0.15506	141	0.32	005.05	0
005	Ash (%)	2009	11.987	0.06600	11.863	0.37506	0.15506	141	0.33	005.05	0
005	Ash (%)	0563	11.994	0.17180	11.863	0.37506	0.15506	141	0.35	005.00	0
005	Ash (%)	0265	11.995	0.01000	11.863	0.37506	0.15506	141	0.35	005.00	0
005	Ash (%)	0643	11.995	0.01000	11.863	0.37506	0.15506	141	0.35	005.00	0
005	Ash (%)	0861	12.000	0.04000	11.863	0.37506	0.15506	141	0.36	005.00	0
005	Ash (%)	0692	12.000	0.00000	11.863	0.37506	0.15506	141	0.36	005.05	0
005	Ash (%)	0353	12.005	0.07000	11.863	0.37506	0.15506	141	0.38	005.00	0
005	Ash (%)	0187	12.010	0.08000	11.863	0.37506	0.15506	141	0.39	005.00	0
005	Ash (%)	2150	12.025	0.23000	11.863	0.37506	0.15506	141	0.43	005.00	0
005	Ash (%)	2006	12.030	0.02000	11.863	0.37506	0.15506	141	0.44	005.00	0
005	Ash (%)	0880	12.030	0.22200	11.863	0.37506	0.15506	141	0.44	005.00	0
005	Ash (%)	0849	12.036	0.02900	11.863	0.37506	0.15506	141	0.46	005.05	0
005	Ash (%)	2075	12.040	0.04000	11.863	0.37506	0.15506	141	0.47	005.02	0
005	Ash (%)	0939	12.045	0.19000	11.863	0.37506	0.15506	141	0.48	005.05	0
005	Ash (%)	0175	12.055	0.13000	11.863	0.37506	0.15506	141	0.51	005.00	0
005	Ash (%)	0646	12.055	0.07000	11.863	0.37506	0.15506	141	0.51	005.00	0
005	Ash (%)	0202	12.055	0.01000	11.863	0.37506	0.15506	141	0.51	005.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0139	12.060	0.12000	11.863	0.37506	0.15506	141	0.52	005.00	0
005	Ash (%)	0722	12.066	0.01530	11.863	0.37506	0.15506	141	0.54	005.00	0
005	Ash (%)	0066	12.070	0.08000	11.863	0.37506	0.15506	141	0.55	005.05	0
005	Ash (%)	0138	12.095	0.03200	11.863	0.37506	0.15506	141	0.62	005.00	0
005	Ash (%)	0183	12.101	0.01900	11.863	0.37506	0.15506	141	0.63	005.00	0
005	Ash (%)	0171	12.105	0.03000	11.863	0.37506	0.15506	141	0.64	005.00	0
005	Ash (%)	0615	12.105	0.03000	11.863	0.37506	0.15506	141	0.64	005.00	0
005	Ash (%)	0171	12.115	0.09000	11.863	0.37506	0.15506	141	0.67	005.05	0
005	Ash (%)	0903	12.125	0.01000	11.863	0.37506	0.15506	141	0.70	005.05	0
005	Ash (%)	0152	12.150	0.10000	11.863	0.37506	0.15506	141	0.76	005.00	0
005	Ash (%)	0366	12.150	0.10000	11.863	0.37506	0.15506	141	0.76	005.00	0
005	Ash (%)	0590	12.150	0.10000	11.863	0.37506	0.15506	141	0.76	005.05	0
005	Ash (%)	0910	12.150	0.10000	11.863	0.37506	0.15506	141	0.76	005.99	0
005	Ash (%)	0884	12.150	0.30000	11.863	0.37506	0.15506	141	0.76	005.99	0
005	Ash (%)	0852	12.165	0.13000	11.863	0.37506	0.15506	141	0.80	005.00	0
005	Ash (%)	2149	12.176	0.01800	11.863	0.37506	0.15506	141	0.83	005.05	0
005	Ash (%)	0345	12.200	0.20000	11.863	0.37506	0.15506	141	0.90	005.00	0
005	Ash (%)	0357	12.200	0.00000	11.863	0.37506	0.15506	141	0.90	005.00	0
005	Ash (%)	0294	12.200	0.00000	11.863	0.37506	0.15506	141	0.90	005.05	0
005	Ash (%)	0610	12.200	0.00000	11.863	0.37506	0.15506	141	0.90	005.05	0
005	Ash (%)	0652	12.200	0.00000	11.863	0.37506	0.15506	141	0.90	005.99	0
005	Ash (%)	0242	12.205	0.05000	11.863	0.37506	0.15506	141	0.91	005.00	0
005	Ash (%)	0689	12.210	0.02000	11.863	0.37506	0.15506	141	0.92	005.05	0
005	Ash (%)	0885	12.235	0.06830	11.863	0.37506	0.15506	141	0.99	005.05	0
005	Ash (%)	0895	12.260	0.00000	11.863	0.37506	0.15506	141	1.06	005.05	0
005	Ash (%)	0889	12.285	0.67000	11.863	0.37506	0.15506	141	1.12	005.05	0
005	Ash (%)	2109	12.285	0.09000	11.863	0.37506	0.15506	141	1.12	005.05	0
005	Ash (%)	0629	12.295	0.01000	11.863	0.37506	0.15506	141	1.15	005.99	0
005	Ash (%)	0675	12.300	0.12000	11.863	0.37506	0.15506	141	1.16	005.00	0
005	Ash (%)	0676	12.305	0.01000	11.863	0.37506	0.15506	141	1.18	005.99	0
005	Ash (%)	0190	12.310	0.06000	11.863	0.37506	0.15506	141	1.19	005.05	0
005	Ash (%)	0581	12.325	0.09000	11.863	0.37506	0.15506	141	1.23	005.00	0
005	Ash (%)	0644	12.345	0.03000	11.863	0.37506	0.15506	141	1.28	005.99	0
005	Ash (%)	0685	12.350	0.00000	11.863	0.37506	0.15506	141	1.30	005.05	0
005	Ash (%)	2089	12.405	0.01000	11.863	0.37506	0.15506	141	1.44	005.00	0
005	Ash (%)	0948	12.463	0.00400	11.863	0.37506	0.15506	141	1.60	005.00	0
005	Ash (%)	2129	12.530	0.02000	11.863	0.37506	0.15506	141	1.78	005.05	0
005	Ash (%)	2120	12.550	0.62000	11.863	0.37506	0.15506	141	1.83	005.11	0
005	Ash (%)	2121	12.705	0.91000	11.863	0.37506	0.15506	141	2.24	005.11	0
005	Ash (%)	0613	12.820	0.42000	11.863	0.37506	0.15506	141	2.55	005.99	0
006	Total sugars (%)	0689	2.3000	0.00000	2.8190	0.41344	0.21000	5	-1.26	006.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
006	Total sugars (%)	0910	2.5000	0.32000	2.8190	0.41344	0.21000	5	-0.77	006.00	0
006	Total sugars (%)	0921	2.8700	0.34000	2.8190	0.41344	0.21000	5	0.12	006.00	0
006	Total sugars (%)	0956	3.2000	0.20000	2.8190	0.41344	0.21000	5	0.92	006.99	0
006	Total sugars (%)	0407	3.2250	0.19000	2.8190	0.41344	0.21000	5	0.98	006.01	0
008	Fiber (%)	0405	3.7400	0.04000	6.9745	0.85489	0.26636	61	-3.78	008.02	0
008	Fiber (%)	2022	4.2200	0.00000	6.9745	0.85489	0.26636	61	-3.22	008.08	0
008	Fiber (%)	0619	5.3800	0.26000	6.9745	0.85489	0.26636	61	-1.87	008.08	0
008	Fiber (%)	0968	5.6875	0.13500	6.9745	0.85489	0.26636	61	-1.51	008.08	0
008	Fiber (%)	0889	5.7200	0.38000	6.9745	0.85489	0.26636	61	-1.47	008.08	0
008	Fiber (%)	0848	5.8050	0.03000	6.9745	0.85489	0.26636	61	-1.37	008.08	0
008	Fiber (%)	0016	5.8950	0.37000	6.9745	0.85489	0.26636	61	-1.26	008.08	0
008	Fiber (%)	0294	6.0500	0.10000	6.9745	0.85489	0.26636	61	-1.08	008.08	0
008	Fiber (%)	0511	6.1000	0.00000	6.9745	0.85489	0.26636	61	-1.02	008.99	0
008	Fiber (%)	0035	6.1100	0.10000	6.9745	0.85489	0.26636	61	-1.01	008.02	0
008	Fiber (%)	0354	6.1600	0.04000	6.9745	0.85489	0.26636	61	-0.95	008.08	0
008	Fiber (%)	0910	6.1600	0.24000	6.9745	0.85489	0.26636	61	-0.95	008.08	0
008	Fiber (%)	0413	6.3000	0.40000	6.9745	0.85489	0.26636	61	-0.79	008.08	0
008	Fiber (%)	0066	6.4050	0.19000	6.9745	0.85489	0.26636	61	-0.67	008.08	0
008	Fiber (%)	0033	6.4150	0.39000	6.9745	0.85489	0.26636	61	-0.65	008.08	0
008	Fiber (%)	0948	6.4950	0.33000	6.9745	0.85489	0.26636	61	-0.56	008.02	0
008	Fiber (%)	0208	6.5050	0.87000	6.9745	0.85489	0.26636	61	-0.55	008.02	0
008	Fiber (%)	0098	6.5100	0.06000	6.9745	0.85489	0.26636	61	-0.54	008.02	0
008	Fiber (%)	2012	6.5250	0.31000	6.9745	0.85489	0.26636	61	-0.53	008.08	0
008	Fiber (%)	0939	6.6150	0.19000	6.9745	0.85489	0.26636	61	-0.42	008.02	0
008	Fiber (%)	0646	6.6900	0.24000	6.9745	0.85489	0.26636	61	-0.33	008.08	0
008	Fiber (%)	0529	6.7000	0.02000	6.9745	0.85489	0.26636	61	-0.32	008.08	0
008	Fiber (%)	0613	6.7250	0.45000	6.9745	0.85489	0.26636	61	-0.29	008.08	0
008	Fiber (%)	0049	6.7950	0.01000	6.9745	0.85489	0.26636	61	-0.21	008.08	0
008	Fiber (%)	0675	6.8050	0.07000	6.9745	0.85489	0.26636	61	-0.20	008.02	0
008	Fiber (%)	0861	6.8150	0.21000	6.9745	0.85489	0.26636	61	-0.19	008.08	0
008	Fiber (%)	0504	6.8400	0.64000	6.9745	0.85489	0.26636	61	-0.16	008.02	0
008	Fiber (%)	0357	6.8500	0.10000	6.9745	0.85489	0.26636	61	-0.15	008.08	0
008	Fiber (%)	0004	6.8800	0.36000	6.9745	0.85489	0.26636	61	-0.11	008.08	0
008	Fiber (%)	0353	6.9100	0.14000	6.9745	0.85489	0.26636	61	-0.08	008.02	0
008	Fiber (%)	0676	6.9200	0.30000	6.9745	0.85489	0.26636	61	-0.06	008.02	0
008	Fiber (%)	0148	7.0250	0.31000	6.9745	0.85489	0.26636	61	0.06	008.08	0
008	Fiber (%)	0693	7.1100	0.08000	6.9745	0.85489	0.26636	61	0.16	008.08	0
008	Fiber (%)	0083	7.1300	0.12000	6.9745	0.85489	0.26636	61	0.18	008.08	0
008	Fiber (%)	0345	7.1500	0.30000	6.9745	0.85489	0.26636	61	0.21	008.02	0
008	Fiber (%)	2129	7.1500	0.32000	6.9745	0.85489	0.26636	61	0.21	008.08	0
008	Fiber (%)	0263	7.1730	0.02000	6.9745	0.85489	0.26636	61	0.23	008.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
008	Fiber (%)	0885	7.2032	0.39130	6.9745	0.85489	0.26636	61	0.27	008.02	0
008	Fiber (%)	0581	7.2400	0.08000	6.9745	0.85489	0.26636	61	0.31	008.08	0
008	Fiber (%)	0190	7.2450	0.43000	6.9745	0.85489	0.26636	61	0.32	008.08	0
008	Fiber (%)	0726	7.2450	0.09000	6.9745	0.85489	0.26636	61	0.32	008.08	0
008	Fiber (%)	0164	7.2500	0.30000	6.9745	0.85489	0.26636	61	0.32	008.08	0
008	Fiber (%)	0148	7.2650	0.01000	6.9745	0.85489	0.26636	61	0.34	008.02	0
008	Fiber (%)	0366	7.4000	0.80000	6.9745	0.85489	0.26636	61	0.50	008.08	0
008	Fiber (%)	0001	7.4200	0.04000	6.9745	0.85489	0.26636	61	0.52	008.08	0
008	Fiber (%)	0037	7.4350	0.31000	6.9745	0.85489	0.26636	61	0.54	008.08	0
008	Fiber (%)	0015	7.5250	0.05000	6.9745	0.85489	0.26636	61	0.64	008.08	0
008	Fiber (%)	0610	7.5500	0.10000	6.9745	0.85489	0.26636	61	0.67	008.99	0
008	Fiber (%)	0358	7.5550	0.61000	6.9745	0.85489	0.26636	61	0.68	008.08	0
008	Fiber (%)	0407	7.5728	0.19690	6.9745	0.85489	0.26636	61	0.70	008.08	0
008	Fiber (%)	0956	7.7500	0.70000	6.9745	0.85489	0.26636	61	0.91	008.08	0
008	Fiber (%)	0226	7.9000	0.00000	6.9745	0.85489	0.26636	61	1.08	008.02	0
008	Fiber (%)	0265	8.0000	0.40000	6.9745	0.85489	0.26636	61	1.20	008.05	0
008	Fiber (%)	0026	8.0700	0.18000	6.9745	0.85489	0.26636	61	1.28	008.08	0
008	Fiber (%)	0045	8.0800	0.64000	6.9745	0.85489	0.26636	61	1.29	008.08	0
008	Fiber (%)	0904	8.2500	0.94000	6.9745	0.85489	0.26636	61	1.49	008.08	0
008	Fiber (%)	0278	8.4500	0.10000	6.9745	0.85489	0.26636	61	1.73	008.08	0
008	Fiber (%)	0536	8.4550	0.59000	6.9745	0.85489	0.26636	61	1.73	008.08	0
008	Fiber (%)	0918	8.4953	0.67490	6.9745	0.85489	0.26636	61	1.78	008.08	0
008	Fiber (%)	0187	8.7250	0.19000	6.9745	0.85489	0.26636	61	2.05	008.02	0
008	Fiber (%)	0510	9.4500	0.30000	6.9745	0.85489	0.26636	61	2.90	008.08	0
009	Fiber (%)	0910	12.850	0.10000	15.391	1.1572	0.43543	49	-2.20	009.07	0
009	Fiber (%)	0413	13.750	0.90000	15.391	1.1572	0.43543	49	-1.42	009.09	0
009	Fiber (%)	0598	14.000	0.20000	15.391	1.1572	0.43543	49	-1.20	009.09	0
009	Fiber (%)	0164	14.050	0.10000	15.391	1.1572	0.43543	49	-1.16	009.09	0
009	Fiber (%)	0353	14.095	0.33000	15.391	1.1572	0.43543	49	-1.12	009.07	0
009	Fiber (%)	0098	14.150	0.90000	15.391	1.1572	0.43543	49	-1.07	009.07	0
009	Fiber (%)	0148	14.205	0.23000	15.391	1.1572	0.43543	49	-1.03	009.09	0
009	Fiber (%)	0049	14.210	0.30000	15.391	1.1572	0.43543	49	-1.02	009.09	0
009	Fiber (%)	0861	14.240	0.38000	15.391	1.1572	0.43543	49	-0.99	009.09	0
009	Fiber (%)	0354	14.345	0.03000	15.391	1.1572	0.43543	49	-0.90	009.09	0
009	Fiber (%)	0889	14.345	0.13000	15.391	1.1572	0.43543	49	-0.90	009.09	0
009	Fiber (%)	0294	14.350	0.10000	15.391	1.1572	0.43543	49	-0.90	009.09	0
009	Fiber (%)	0581	14.385	0.15000	15.391	1.1572	0.43543	49	-0.87	009.09	0
009	Fiber (%)	0939	14.525	0.13000	15.391	1.1572	0.43543	49	-0.75	009.09	0
009	Fiber (%)	0263	14.598	0.02340	15.391	1.1572	0.43543	49	-0.69	009.09	0
009	Fiber (%)	0693	14.600	0.00000	15.391	1.1572	0.43543	49	-0.68	009.09	0
009	Fiber (%)	0504	14.620	0.50000	15.391	1.1572	0.43543	49	-0.67	009.04	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
009	Fiber (%)	0613	14.900	1.0200	15.391	1.1572	0.43543	49	-0.42	009.09	0
009	Fiber (%)	0358	15.015	0.59000	15.391	1.1572	0.43543	49	-0.33	009.09	0
009	Fiber (%)	0037	15.100	0.12000	15.391	1.1572	0.43543	49	-0.25	009.09	0
009	Fiber (%)	2022	15.120	0.04000	15.391	1.1572	0.43543	49	-0.23	009.09	0
009	Fiber (%)	0066	15.130	0.18000	15.391	1.1572	0.43543	49	-0.23	009.09	0
009	Fiber (%)	0956	15.350	0.10000	15.391	1.1572	0.43543	49	-0.04	009.09	0
009	Fiber (%)	0675	15.415	0.07000	15.391	1.1572	0.43543	49	0.02	009.07	0
009	Fiber (%)	0083	15.420	0.40000	15.391	1.1572	0.43543	49	0.02	009.09	0
009	Fiber (%)	0510	15.500	0.00000	15.391	1.1572	0.43543	49	0.09	009.09	0
009	Fiber (%)	2012	15.545	0.19000	15.391	1.1572	0.43543	49	0.13	009.09	0
009	Fiber (%)	0366	15.550	1.1000	15.391	1.1572	0.43543	49	0.14	009.09	0
009	Fiber (%)	0619	15.650	0.10000	15.391	1.1572	0.43543	49	0.22	009.09	0
009	Fiber (%)	0646	15.710	0.22000	15.391	1.1572	0.43543	49	0.28	009.09	0
009	Fiber (%)	0045	15.750	0.70000	15.391	1.1572	0.43543	49	0.31	009.09	0
009	Fiber (%)	0948	15.800	0.18000	15.391	1.1572	0.43543	49	0.35	009.09	0
009	Fiber (%)	0918	15.881	0.61090	15.391	1.1572	0.43543	49	0.42	009.09	0
009	Fiber (%)	0610	16.000	0.40000	15.391	1.1572	0.43543	49	0.53	009.07	0
009	Fiber (%)	0015	16.000	0.40000	15.391	1.1572	0.43543	49	0.53	009.09	0
009	Fiber (%)	0278	16.050	0.10000	15.391	1.1572	0.43543	49	0.57	009.09	0
009	Fiber (%)	0345	16.100	0.20000	15.391	1.1572	0.43543	49	0.61	009.07	0
009	Fiber (%)	0226	16.450	0.10000	15.391	1.1572	0.43543	49	0.92	009.07	0
009	Fiber (%)	0265	16.450	0.30000	15.391	1.1572	0.43543	49	0.92	009.09	0
009	Fiber (%)	0407	16.505	0.48080	15.391	1.1572	0.43543	49	0.96	009.09	0
009	Fiber (%)	0968	16.507	0.54100	15.391	1.1572	0.43543	49	0.96	009.09	0
009	Fiber (%)	0536	16.510	0.36000	15.391	1.1572	0.43543	49	0.97	009.09	0
009	Fiber (%)	0357	16.600	1.0000	15.391	1.1572	0.43543	49	1.04	009.09	0
009	Fiber (%)	0904	16.750	0.26000	15.391	1.1572	0.43543	49	1.17	009.09	0
009	Fiber (%)	0187	16.775	0.09000	15.391	1.1572	0.43543	49	1.20	009.07	0
009	Fiber (%)	0889	17.180	0.04000	15.391	1.1572	0.43543	49	1.55	009.99	0
009	Fiber (%)	0676	17.950	0.32000	15.391	1.1572	0.43543	49	2.21	009.07	0
009	Fiber (%)	0613	18.285	6.5500	15.391	1.1572	0.43543	49	2.50	009.99	0
009	Fiber (%)	0001	18.585	0.07000	15.391	1.1572	0.43543	49	2.76	009.09	0
010	Moisture (%)	0017	7.1000	0.20000	9.5547	0.61199	0.16692	25	-4.01	010.99	0
010	Moisture (%)	0722	8.7186	0.03890	9.5547	0.61199	0.16692	25	-1.37	010.99	0
010	Moisture (%)	0843	8.7750	0.71000	9.5547	0.61199	0.16692	25	-1.27	010.03	0
010	Moisture (%)	2121	9.0650	0.03000	9.5547	0.61199	0.16692	25	-0.80	010.11	0
010	Moisture (%)	2089	9.0700	0.04000	9.5547	0.61199	0.16692	25	-0.79	010.99	0
010	Moisture (%)	2120	9.1000	0.18000	9.5547	0.61199	0.16692	25	-0.74	010.11	0
010	Moisture (%)	0948	9.2300	0.14000	9.5547	0.61199	0.16692	25	-0.53	010.99	0
010	Moisture (%)	0917	9.3000	0.20000	9.5547	0.61199	0.16692	25	-0.42	010.99	0
010	Moisture (%)	0952	9.3600	0.00000	9.5547	0.61199	0.16692	25	-0.32	010.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
010	Moisture (%)	0164	9.3800	0.54000	9.5547	0.61199	0.16692	25	-0.29	010.03	0
010	Moisture (%)	0939	9.4050	0.07000	9.5547	0.61199	0.16692	25	-0.24	010.99	0
010	Moisture (%)	2129	9.4150	0.19000	9.5547	0.61199	0.16692	25	-0.23	010.99	0
010	Moisture (%)	0692	9.4500	0.30000	9.5547	0.61199	0.16692	25	-0.17	010.99	0
010	Moisture (%)	0208	9.5550	0.25000	9.5547	0.61199	0.16692	25	0.00	010.03	0
010	Moisture (%)	0726	9.6300	0.10000	9.5547	0.61199	0.16692	25	0.12	010.99	0
010	Moisture (%)	0190	9.6650	0.01000	9.5547	0.61199	0.16692	25	0.18	010.99	0
010	Moisture (%)	2150	9.6700	0.20000	9.5547	0.61199	0.16692	25	0.19	010.99	0
010	Moisture (%)	0921	9.7300	0.10000	9.5547	0.61199	0.16692	25	0.29	010.99	0
010	Moisture (%)	0884	9.7500	0.10000	9.5547	0.61199	0.16692	25	0.32	010.99	0
010	Moisture (%)	0889	9.9850	0.09000	9.5547	0.61199	0.16692	25	0.70	010.11	0
010	Moisture (%)	2127	10.240	0.24000	9.5547	0.61199	0.16692	25	1.12	010.99	0
010	Moisture (%)	2128	10.319	0.03400	9.5547	0.61199	0.16692	25	1.25	010.11	0
010	Moisture (%)	0553	10.505	0.33000	9.5547	0.61199	0.16692	25	1.55	010.11	0
010	Moisture (%)	0852	10.740	0.06000	9.5547	0.61199	0.16692	25	1.94	010.11	0
010	Moisture (%)	0405	10.740	0.02000	9.5547	0.61199	0.16692	25	1.94	010.99	0
011	Loss on Drying (%)	0874	8.3650	0.25000	10.132	0.36175	0.11703	71	-4.88	011.01	0
011	Loss on Drying (%)	1019	8.4400	0.06000	10.132	0.36175	0.11703	71	-4.68	011.01	0
011	Loss on Drying (%)	2031	8.8050	0.73000	10.132	0.36175	0.11703	71	-3.67	011.01	0
011	Loss on Drying (%)	0536	8.9100	0.10000	10.132	0.36175	0.11703	71	-3.38	011.01	0
011	Loss on Drying (%)	0345	8.9200	0.04000	10.132	0.36175	0.11703	71	-3.35	011.01	0
011	Loss on Drying (%)	2075	9.5100	0.04000	10.132	0.36175	0.11703	71	-1.72	011.01	0
011	Loss on Drying (%)	0643	9.5500	0.10000	10.132	0.36175	0.11703	71	-1.61	011.01	0
011	Loss on Drying (%)	0852	9.5800	0.12000	10.132	0.36175	0.11703	71	-1.53	011.99	0
011	Loss on Drying (%)	0552	9.5900	0.22000	10.132	0.36175	0.11703	71	-1.50	011.01	0
011	Loss on Drying (%)	0674	9.6000	0.00000	10.132	0.36175	0.11703	71	-1.47	011.01	0
011	Loss on Drying (%)	0563	9.6658	0.01930	10.132	0.36175	0.11703	71	-1.29	011.01	0
011	Loss on Drying (%)	0175	9.7000	0.00000	10.132	0.36175	0.11703	71	-1.19	011.01	0
011	Loss on Drying (%)	0857	9.7500	0.10000	10.132	0.36175	0.11703	71	-1.06	011.99	0
011	Loss on Drying (%)	0598	9.7700	0.14000	10.132	0.36175	0.11703	71	-1.00	011.01	0
011	Loss on Drying (%)	2076	9.7800	0.12000	10.132	0.36175	0.11703	71	-0.97	011.99	0
011	Loss on Drying (%)	0953	9.8400	0.04000	10.132	0.36175	0.11703	71	-0.81	011.01	0
011	Loss on Drying (%)	0026	9.8450	0.45000	10.132	0.36175	0.11703	71	-0.79	011.01	0
011	Loss on Drying (%)	0646	9.8750	0.03000	10.132	0.36175	0.11703	71	-0.71	011.01	0
011	Loss on Drying (%)	0148	9.9650	0.57000	10.132	0.36175	0.11703	71	-0.46	011.01	0
011	Loss on Drying (%)	0722	9.9922	0.07340	10.132	0.36175	0.11703	71	-0.39	011.01	0
011	Loss on Drying (%)	0138	9.9965	0.02900	10.132	0.36175	0.11703	71	-0.37	011.01	0
011	Loss on Drying (%)	2006	10.025	0.05000	10.132	0.36175	0.11703	71	-0.30	011.01	0
011	Loss on Drying (%)	0574	10.040	0.08000	10.132	0.36175	0.11703	71	-0.25	011.01	0
011	Loss on Drying (%)	0541	10.045	0.13000	10.132	0.36175	0.11703	71	-0.24	011.99	0
011	Loss on Drying (%)	0298	10.080	0.02000	10.132	0.36175	0.11703	71	-0.14	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
011	Loss on Drying (%)	0425	10.080	0.04000	10.132	0.36175	0.11703	71	-0.14	011.02	0
011	Loss on Drying (%)	0202	10.090	0.02000	10.132	0.36175	0.11703	71	-0.12	011.01	0
011	Loss on Drying (%)	0100	10.095	0.25000	10.132	0.36175	0.11703	71	-0.10	011.01	0
011	Loss on Drying (%)	0171	10.100	0.04000	10.132	0.36175	0.11703	71	-0.09	011.01	0
011	Loss on Drying (%)	0294	10.100	0.00000	10.132	0.36175	0.11703	71	-0.09	011.01	0
011	Loss on Drying (%)	0066	10.100	0.10000	10.132	0.36175	0.11703	71	-0.09	011.01	0
011	Loss on Drying (%)	0226	10.110	0.04000	10.132	0.36175	0.11703	71	-0.06	011.01	0
011	Loss on Drying (%)	0942	10.140	0.20000	10.132	0.36175	0.11703	71	0.02	011.02	0
011	Loss on Drying (%)	0650	10.150	0.44000	10.132	0.36175	0.11703	71	0.05	011.01	0
011	Loss on Drying (%)	0263	10.150	0.00000	10.132	0.36175	0.11703	71	0.05	011.01	0
011	Loss on Drying (%)	0859	10.171	0.01500	10.132	0.36175	0.11703	71	0.11	011.01	0
011	Loss on Drying (%)	0675	10.175	0.01000	10.132	0.36175	0.11703	71	0.12	011.01	0
011	Loss on Drying (%)	0520	10.215	0.37000	10.132	0.36175	0.11703	71	0.23	011.01	0
011	Loss on Drying (%)	0653	10.230	0.02000	10.132	0.36175	0.11703	71	0.27	011.01	0
011	Loss on Drying (%)	2073	10.230	0.30000	10.132	0.36175	0.11703	71	0.27	011.01	0
011	Loss on Drying (%)	0144	10.250	0.16000	10.132	0.36175	0.11703	71	0.33	011.01	0
011	Loss on Drying (%)	0208	10.250	0.10000	10.132	0.36175	0.11703	71	0.33	011.01	0
011	Loss on Drying (%)	0880	10.259	0.34100	10.132	0.36175	0.11703	71	0.35	011.01	0
011	Loss on Drying (%)	0861	10.265	0.07000	10.132	0.36175	0.11703	71	0.37	011.01	0
011	Loss on Drying (%)	0229	10.300	0.00000	10.132	0.36175	0.11703	71	0.46	011.01	0
011	Loss on Drying (%)	0510	10.300	0.00000	10.132	0.36175	0.11703	71	0.46	011.01	0
011	Loss on Drying (%)	0682	10.300	0.00000	10.132	0.36175	0.11703	71	0.46	011.01	0
011	Loss on Drying (%)	0620	10.304	0.10590	10.132	0.36175	0.11703	71	0.48	011.01	0
011	Loss on Drying (%)	0539	10.305	0.37000	10.132	0.36175	0.11703	71	0.48	011.01	0
011	Loss on Drying (%)	0511	10.315	0.05000	10.132	0.36175	0.11703	71	0.51	011.01	0
011	Loss on Drying (%)	0685	10.325	0.01000	10.132	0.36175	0.11703	71	0.53	011.01	0
011	Loss on Drying (%)	0848	10.335	0.01000	10.132	0.36175	0.11703	71	0.56	011.01	0
011	Loss on Drying (%)	0164	10.340	0.04000	10.132	0.36175	0.11703	71	0.58	011.01	0
011	Loss on Drying (%)	0242	10.395	0.15000	10.132	0.36175	0.11703	71	0.73	011.01	0
011	Loss on Drying (%)	0033	10.400	0.00000	10.132	0.36175	0.11703	71	0.74	011.01	0
011	Loss on Drying (%)	0152	10.400	0.00000	10.132	0.36175	0.11703	71	0.74	011.01	0
011	Loss on Drying (%)	0265	10.400	0.00000	10.132	0.36175	0.11703	71	0.74	011.01	0
011	Loss on Drying (%)	0843	10.410	0.40000	10.132	0.36175	0.11703	71	0.77	011.01	0
011	Loss on Drying (%)	0358	10.480	0.02000	10.132	0.36175	0.11703	71	0.96	011.01	0
011	Loss on Drying (%)	0553	10.480	0.16000	10.132	0.36175	0.11703	71	0.96	011.01	0
011	Loss on Drying (%)	0589	10.480	0.14000	10.132	0.36175	0.11703	71	0.96	011.01	0
011	Loss on Drying (%)	0876	10.480	0.06000	10.132	0.36175	0.11703	71	0.96	011.01	0
011	Loss on Drying (%)	0354	10.490	0.00000	10.132	0.36175	0.11703	71	0.99	011.01	0
011	Loss on Drying (%)	0541	10.505	0.01000	10.132	0.36175	0.11703	71	1.03	011.01	0
011	Loss on Drying (%)	0108	10.525	0.07000	10.132	0.36175	0.11703	71	1.09	011.01	0
011	Loss on Drying (%)	0098	10.550	0.08000	10.132	0.36175	0.11703	71	1.16	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
011	Loss on Drying (%)	0205	10.550	0.10000	10.132	0.36175	0.11703	71	1.16	011.01	0
011	Loss on Drying (%)	0968	10.568	0.03200	10.132	0.36175	0.11703	71	1.21	011.01	0
011	Loss on Drying (%)	0233	10.640	0.26000	10.132	0.36175	0.11703	71	1.40	011.01	0
011	Loss on Drying (%)	2022	10.690	0.08000	10.132	0.36175	0.11703	71	1.54	011.01	0
011	Loss on Drying (%)	0407	10.751	0.13360	10.132	0.36175	0.11703	71	1.71	011.01	0
012	Starch (%)	0918	22.423	0.66000	29.491	1.5566	0.40599	32	-4.54	012.01	0
012	Starch (%)	0613	26.225	2.3900	29.491	1.5566	0.40599	32	-2.10	012.01	0
012	Starch (%)	0676	26.660	0.06000	29.491	1.5566	0.40599	32	-1.82	012.01	0
012	Starch (%)	0265	27.150	0.30000	29.491	1.5566	0.40599	32	-1.50	012.01	0
012	Starch (%)	0164	27.450	0.50000	29.491	1.5566	0.40599	32	-1.31	012.01	0
012	Starch (%)	0160	28.340	0.82000	29.491	1.5566	0.40599	32	-0.74	012.04	0
012	Starch (%)	0693	28.350	0.90000	29.491	1.5566	0.40599	32	-0.73	012.01	0
012	Starch (%)	0956	28.800	2.0000	29.491	1.5566	0.40599	32	-0.44	012.01	0
012	Starch (%)	0278	28.950	0.10000	29.491	1.5566	0.40599	32	-0.35	012.04	0
012	Starch (%)	0510	29.000	0.20000	29.491	1.5566	0.40599	32	-0.32	012.04	0
012	Starch (%)	0889	29.005	0.01000	29.491	1.5566	0.40599	32	-0.31	012.03	0
012	Starch (%)	0885	29.019	0.15850	29.491	1.5566	0.40599	32	-0.30	012.03	0
012	Starch (%)	0353	29.040	0.16000	29.491	1.5566	0.40599	32	-0.29	012.04	0
012	Starch (%)	0148	29.050	0.16000	29.491	1.5566	0.40599	32	-0.28	012.03	0
012	Starch (%)	0098	29.100	0.20000	29.491	1.5566	0.40599	32	-0.25	012.03	0
012	Starch (%)	0366	29.200	0.00000	29.491	1.5566	0.40599	32	-0.19	012.01	0
012	Starch (%)	0045	29.250	0.30000	29.491	1.5566	0.40599	32	-0.15	012.01	0
012	Starch (%)	0948	29.980	0.16000	29.491	1.5566	0.40599	32	0.31	012.04	0
012	Starch (%)	0164	30.000	0.80000	29.491	1.5566	0.40599	32	0.33	012.00	0
012	Starch (%)	0689	30.100	0.40000	29.491	1.5566	0.40599	32	0.39	012.00	0
012	Starch (%)	0955	30.150	0.10000	29.491	1.5566	0.40599	32	0.42	012.02	0
012	Starch (%)	0407	30.170	0.52000	29.491	1.5566	0.40599	32	0.44	012.03	0
012	Starch (%)	0552	30.275	0.11000	29.491	1.5566	0.40599	32	0.50	012.00	0
012	Starch (%)	2006	30.310	0.16000	29.491	1.5566	0.40599	32	0.53	012.00	0
012	Starch (%)	0889	30.665	0.55000	29.491	1.5566	0.40599	32	0.75	012.11	0
012	Starch (%)	0852	30.780	0.14000	29.491	1.5566	0.40599	32	0.83	012.11	0
012	Starch (%)	0354	30.935	0.01000	29.491	1.5566	0.40599	32	0.93	012.00	0
012	Starch (%)	0910	31.000	0.60000	29.491	1.5566	0.40599	32	0.97	012.00	0
012	Starch (%)	2009	31.148	0.12120	29.491	1.5566	0.40599	32	1.06	012.00	0
012	Starch (%)	0619	31.375	0.13000	29.491	1.5566	0.40599	32	1.21	012.00	0
012	Starch (%)	2128	32.394	0.00200	29.491	1.5566	0.40599	32	1.86	012.11	0
012	Starch (%)	2089	33.665	0.27000	29.491	1.5566	0.40599	32	2.68	012.02	0
013	Fat (%)	0027	4.4820	0.44800	6.1557	0.52717	0.19170	51	-3.17	013.13	0
013	Fat (%)	0618	4.6850	0.25000	6.1557	0.52717	0.19170	51	-2.79	013.08	0
013	Fat (%)	0618	5.1200	0.02000	6.1557	0.52717	0.19170	51	-1.96	013.00	0
013	Fat (%)	0036	5.1475	0.02730	6.1557	0.52717	0.19170	51	-1.91	013.10	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
013	Fat (%)	0689	5.2500	0.10000	6.1557	0.52717	0.19170	51	-1.72	013.00	0
013	Fat (%)	0083	5.3200	0.04000	6.1557	0.52717	0.19170	51	-1.59	013.00	0
013	Fat (%)	0504	5.3800	0.26000	6.1557	0.52717	0.19170	51	-1.47	013.00	0
013	Fat (%)	0017	5.4500	0.30000	6.1557	0.52717	0.19170	51	-1.34	013.02	0
013	Fat (%)	0968	5.7195	0.13700	6.1557	0.52717	0.19170	51	-0.83	013.13	0
013	Fat (%)	0541	5.7700	0.46000	6.1557	0.52717	0.19170	51	-0.73	013.00	0
013	Fat (%)	2128	5.8576	0.06460	6.1557	0.52717	0.19170	51	-0.57	013.12	0
013	Fat (%)	0171	5.8650	0.21000	6.1557	0.52717	0.19170	51	-0.55	013.02	0
013	Fat (%)	0610	5.8800	0.28000	6.1557	0.52717	0.19170	51	-0.52	013.10	0
013	Fat (%)	0904	5.8950	0.09000	6.1557	0.52717	0.19170	51	-0.49	013.13	0
013	Fat (%)	0843	5.9200	0.34000	6.1557	0.52717	0.19170	51	-0.45	013.02	0
013	Fat (%)	0098	5.9300	0.10000	6.1557	0.52717	0.19170	51	-0.43	013.02	0
013	Fat (%)	2073	5.9400	0.14000	6.1557	0.52717	0.19170	51	-0.41	013.02	0
013	Fat (%)	0884	5.9500	0.10000	6.1557	0.52717	0.19170	51	-0.39	013.00	0
013	Fat (%)	0910	5.9850	0.27000	6.1557	0.52717	0.19170	51	-0.32	013.00	0
013	Fat (%)	0629	6.0300	0.00000	6.1557	0.52717	0.19170	51	-0.24	013.10	0
013	Fat (%)	0948	6.0500	0.06000	6.1557	0.52717	0.19170	51	-0.20	013.00	0
013	Fat (%)	2006	6.0700	0.74000	6.1557	0.52717	0.19170	51	-0.16	013.10	0
013	Fat (%)	0650	6.0750	0.75000	6.1557	0.52717	0.19170	51	-0.15	013.02	0
013	Fat (%)	0652	6.1000	0.00000	6.1557	0.52717	0.19170	51	-0.11	013.00	0
013	Fat (%)	0242	6.1050	0.27000	6.1557	0.52717	0.19170	51	-0.10	013.00	0
013	Fat (%)	0353	6.1600	0.08000	6.1557	0.52717	0.19170	51	0.01	013.10	0
013	Fat (%)	0208	6.1900	0.20000	6.1557	0.52717	0.19170	51	0.07	013.02	0
013	Fat (%)	0229	6.2000	0.10000	6.1557	0.52717	0.19170	51	0.08	013.00	0
013	Fat (%)	0581	6.2350	0.25000	6.1557	0.52717	0.19170	51	0.15	013.13	0
013	Fat (%)	0643	6.3150	0.07000	6.1557	0.52717	0.19170	51	0.30	013.02	0
013	Fat (%)	0874	6.3250	0.49000	6.1557	0.52717	0.19170	51	0.32	013.02	0
013	Fat (%)	0553	6.3550	0.33000	6.1557	0.52717	0.19170	51	0.38	013.02	0
013	Fat (%)	0074	6.3800	0.08000	6.1557	0.52717	0.19170	51	0.43	013.02	0
013	Fat (%)	0955	6.4050	0.05000	6.1557	0.52717	0.19170	51	0.47	013.13	0
013	Fat (%)	0100	6.4200	0.00000	6.1557	0.52717	0.19170	51	0.50	013.02	0
013	Fat (%)	0921	6.4700	0.12000	6.1557	0.52717	0.19170	51	0.60	013.02	0
013	Fat (%)	0354	6.5050	0.05000	6.1557	0.52717	0.19170	51	0.66	013.02	0
013	Fat (%)	0615	6.5100	0.02000	6.1557	0.52717	0.19170	51	0.67	013.00	0
013	Fat (%)	0148	6.5150	0.11000	6.1557	0.52717	0.19170	51	0.68	013.02	0
013	Fat (%)	0164	6.5150	0.11000	6.1557	0.52717	0.19170	51	0.68	013.02	0
013	Fat (%)	2150	6.5800	0.16000	6.1557	0.52717	0.19170	51	0.80	013.02	0
013	Fat (%)	2057	6.5950	0.05000	6.1557	0.52717	0.19170	51	0.83	013.00	0
013	Fat (%)	0202	6.6100	0.26000	6.1557	0.52717	0.19170	51	0.86	013.00	0
013	Fat (%)	2076	6.6500	0.76000	6.1557	0.52717	0.19170	51	0.94	013.00	0
013	Fat (%)	0010	6.7000	0.40000	6.1557	0.52717	0.19170	51	1.03	013.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
013	Fat (%)	0682	6.7000	0.00000	6.1557	0.52717	0.19170	51	1.03	013.02	0
013	Fat (%)	0139	6.7400	0.12000	6.1557	0.52717	0.19170	51	1.11	013.00	0
013	Fat (%)	0676	6.8650	0.01000	6.1557	0.52717	0.19170	51	1.35	013.02	0
013	Fat (%)	0675	6.9600	0.18000	6.1557	0.52717	0.19170	51	1.53	013.02	0
013	Fat (%)	0265	7.0500	0.30000	6.1557	0.52717	0.19170	51	1.70	013.13	0
013	Fat (%)	0861	8.2500	0.02000	6.1557	0.52717	0.19170	51	3.97	013.02	0
015	Aluminum (mg / kg (ppm))	0049	85.025	14.610	154.50	47.081	18.716	12	-1.48	015.41	0
015	Aluminum (mg / kg (ppm))	0918	107.40	3.5810	154.50	47.081	18.716	12	-1.00	015.43	0
015	Aluminum (mg / kg (ppm))	0520	110.35	3.9000	154.50	47.081	18.716	12	-0.94	015.41	0
015	Aluminum (mg / kg (ppm))	0171	135.00	4.0000	154.50	47.081	18.716	12	-0.41	015.41	0
015	Aluminum (mg / kg (ppm))	0345	142.00	6.0000	154.50	47.081	18.716	12	-0.27	015.43	0
015	Aluminum (mg / kg (ppm))	0510	153.00	4.0000	154.50	47.081	18.716	12	-0.03	015.43	0
015	Aluminum (mg / kg (ppm))	0560	153.75	19.500	154.50	47.081	18.716	12	-0.02	015.42	0
015	Aluminum (mg / kg (ppm))	0164	163.50	3.0000	154.50	47.081	18.716	12	0.19	015.41	0
015	Aluminum (mg / kg (ppm))	0021	185.00	90.000	154.50	47.081	18.716	12	0.65	015.43	0
015	Aluminum (mg / kg (ppm))	0353	194.00	8.0000	154.50	47.081	18.716	12	0.84	015.43	0
015	Aluminum (mg / kg (ppm))	0169	204.50	55.000	154.50	47.081	18.716	12	1.06	015.43	0
015	Aluminum (mg / kg (ppm))	0553	220.50	13.000	154.50	47.081	18.716	12	1.40	015.53	0
017	Boron (mg / kg (ppm))	2053	4.9000	0.08000	6.2948	0.96114	0.37068	15	-1.45	017.42	0
017	Boron (mg / kg (ppm))	0226	5.2650	0.03000	6.2948	0.96114	0.37068	15	-1.07	017.41	0
017	Boron (mg / kg (ppm))	0045	5.5600	0.42000	6.2948	0.96114	0.37068	15	-0.76	017.42	0
017	Boron (mg / kg (ppm))	0345	5.6500	0.10000	6.2948	0.96114	0.37068	15	-0.67	017.43	0
017	Boron (mg / kg (ppm))	0096	5.7785	0.19020	6.2948	0.96114	0.37068	15	-0.54	017.52	0
017	Boron (mg / kg (ppm))	0358	5.8400	0.34000	6.2948	0.96114	0.37068	15	-0.47	017.41	0
017	Boron (mg / kg (ppm))	0693	5.9300	0.00000	6.2948	0.96114	0.37068	15	-0.38	017.42	0
017	Boron (mg / kg (ppm))	0560	6.2900	0.44000	6.2948	0.96114	0.37068	15	0.00	017.42	0
017	Boron (mg / kg (ppm))	0353	6.5000	1.0000	6.2948	0.96114	0.37068	15	0.21	017.43	0
017	Boron (mg / kg (ppm))	0918	6.5000	0.64000	6.2948	0.96114	0.37068	15	0.21	017.43	0
017	Boron (mg / kg (ppm))	0229	6.6500	1.8600	6.2948	0.96114	0.37068	15	0.37	017.41	0
017	Boron (mg / kg (ppm))	0510	7.0000	0.00000	6.2948	0.96114	0.37068	15	0.73	017.43	0
017	Boron (mg / kg (ppm))	0049	7.0850	0.29000	6.2948	0.96114	0.37068	15	0.82	017.41	0
017	Boron (mg / kg (ppm))	0187	7.8850	0.03000	6.2948	0.96114	0.37068	15	1.65	017.42	0
017	Boron (mg / kg (ppm))	0294	7.9300	0.14000	6.2948	0.96114	0.37068	15	1.70	017.42	0
017	Boron (mg / kg (ppm))	0021	0.00000	0.00000	6.2948	0.96114	0.37068	15		017.43	4
019	Calcium (%)	0405	0.42125	0.00170	3.5686	0.18013	4.0374	143	-17.47	019.41	0
019	Calcium (%)	0017	2.8300	0.26000	3.5686	0.18013	4.0374	143	-4.10	019.43	0
019	Calcium (%)	0529	2.8350	0.33000	3.5686	0.18013	4.0374	143	-4.07	019.31	0
019	Calcium (%)	0904	3.1650	0.05000	3.5686	0.18013	4.0374	143	-2.24	019.31	0
019	Calcium (%)	0001	3.1735	0.01100	3.5686	0.18013	4.0374	143	-2.19	019.31	0
019	Calcium (%)	0043	3.1850	0.01000	3.5686	0.18013	4.0374	143	-2.13	019.43	0
019	Calcium (%)	0650	3.2000	0.38000	3.5686	0.18013	4.0374	143	-2.05	019.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	0026	3.2200	0.16000	3.5686	0.18013	4.0374	143	-1.94	019.42	0
019	Calcium (%)	0889	3.2440	0.00400	3.5686	0.18013	4.0374	143	-1.80	019.99	0
019	Calcium (%)	0014	3.2450	0.03000	3.5686	0.18013	4.0374	143	-1.80	019.31	0
019	Calcium (%)	0144	3.2550	0.05000	3.5686	0.18013	4.0374	143	-1.74	019.41	0
019	Calcium (%)	0884	3.2600	0.28000	3.5686	0.18013	4.0374	143	-1.71	019.31	0
019	Calcium (%)	0552	3.2650	0.41000	3.5686	0.18013	4.0374	143	-1.69	019.00	0
019	Calcium (%)	0089	3.2800	0.00000	3.5686	0.18013	4.0374	143	-1.60	019.41	0
019	Calcium (%)	2006	3.3015	0.01300	3.5686	0.18013	4.0374	143	-1.48	019.09	0
019	Calcium (%)	0096	3.3061	0.09250	3.5686	0.18013	4.0374	143	-1.46	019.42	0
019	Calcium (%)	0263	3.3125	0.01500	3.5686	0.18013	4.0374	143	-1.42	019.32	0
019	Calcium (%)	0021	3.3500	0.30000	3.5686	0.18013	4.0374	143	-1.21	019.43	0
019	Calcium (%)	0848	3.3700	0.04000	3.5686	0.18013	4.0374	143	-1.10	019.41	0
019	Calcium (%)	0357	3.4166	0.12110	3.5686	0.18013	4.0374	143	-0.84	019.42	0
019	Calcium (%)	0019	3.4250	0.07000	3.5686	0.18013	4.0374	143	-0.80	019.41	0
019	Calcium (%)	2129	3.4250	0.01000	3.5686	0.18013	4.0374	143	-0.80	019.42	0
019	Calcium (%)	0939	3.4400	0.10000	3.5686	0.18013	4.0374	143	-0.71	019.31	0
019	Calcium (%)	0510	3.4400	0.02000	3.5686	0.18013	4.0374	143	-0.71	019.43	0
019	Calcium (%)	0629	3.4500	0.00000	3.5686	0.18013	4.0374	143	-0.66	019.08	0
019	Calcium (%)	2062	3.4538	0.03460	3.5686	0.18013	4.0374	143	-0.64	019.00	0
019	Calcium (%)	0038	3.4550	0.03000	3.5686	0.18013	4.0374	143	-0.63	019.31	0
019	Calcium (%)	0199	3.4550	0.09000	3.5686	0.18013	4.0374	143	-0.63	019.53	0
019	Calcium (%)	0590	3.4600	0.14000	3.5686	0.18013	4.0374	143	-0.60	019.08	0
019	Calcium (%)	0298	3.4600	0.02000	3.5686	0.18013	4.0374	143	-0.60	019.41	0
019	Calcium (%)	0187	3.4600	0.02000	3.5686	0.18013	4.0374	143	-0.60	019.42	0
019	Calcium (%)	0265	3.4600	0.30000	3.5686	0.18013	4.0374	143	-0.60	019.42	0
019	Calcium (%)	0849	3.4615	0.02300	3.5686	0.18013	4.0374	143	-0.59	019.08	0
019	Calcium (%)	0512	3.4615	0.03900	3.5686	0.18013	4.0374	143	-0.59	019.43	0
019	Calcium (%)	0096	3.4617	0.00980	3.5686	0.18013	4.0374	143	-0.59	019.52	0
019	Calcium (%)	0154	3.4626	0.28530	3.5686	0.18013	4.0374	143	-0.59	019.52	0
019	Calcium (%)	0039	3.4640	0.05000	3.5686	0.18013	4.0374	143	-0.58	019.44	0
019	Calcium (%)	0142	3.4651	0.10320	3.5686	0.18013	4.0374	143	-0.57	019.31	0
019	Calcium (%)	0171	3.4700	0.02000	3.5686	0.18013	4.0374	143	-0.55	019.41	0
019	Calcium (%)	0035	3.4700	0.00000	3.5686	0.18013	4.0374	143	-0.55	019.43	0
019	Calcium (%)	0560	3.4717	0.10050	3.5686	0.18013	4.0374	143	-0.54	019.42	0
019	Calcium (%)	0345	3.4725	0.09500	3.5686	0.18013	4.0374	143	-0.53	019.43	0
019	Calcium (%)	0208	3.4785	0.26700	3.5686	0.18013	4.0374	143	-0.50	019.41	0
019	Calcium (%)	0354	3.4800	0.02000	3.5686	0.18013	4.0374	143	-0.49	019.31	0
019	Calcium (%)	0036	3.4804	0.03260	3.5686	0.18013	4.0374	143	-0.49	019.43	0
019	Calcium (%)	2109	3.4825	0.05300	3.5686	0.18013	4.0374	143	-0.48	019.00	0
019	Calcium (%)	0897	3.4850	0.01000	3.5686	0.18013	4.0374	143	-0.46	019.00	0
019	Calcium (%)	2022	3.4900	0.00000	3.5686	0.18013	4.0374	143	-0.44	019.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	0164	3.4900	0.04000	3.5686	0.18013	4.0374	143	-0.44	019.41	0
019	Calcium (%)	0903	3.4950	0.01000	3.5686	0.18013	4.0374	143	-0.41	019.00	0
019	Calcium (%)	0511	3.4950	0.03000	3.5686	0.18013	4.0374	143	-0.41	019.41	0
019	Calcium (%)	0205	3.4950	0.05000	3.5686	0.18013	4.0374	143	-0.41	019.42	0
019	Calcium (%)	2097	3.4971	0.03420	3.5686	0.18013	4.0374	143	-0.40	019.31	0
019	Calcium (%)	0294	3.5050	0.01000	3.5686	0.18013	4.0374	143	-0.35	019.42	0
019	Calcium (%)	0139	3.5150	0.01000	3.5686	0.18013	4.0374	143	-0.30	019.41	0
019	Calcium (%)	0035	3.5200	0.02000	3.5686	0.18013	4.0374	143	-0.27	019.33	0
019	Calcium (%)	0358	3.5200	0.04000	3.5686	0.18013	4.0374	143	-0.27	019.41	0
019	Calcium (%)	0553	3.5200	0.14000	3.5686	0.18013	4.0374	143	-0.27	019.53	0
019	Calcium (%)	0895	3.5250	0.01000	3.5686	0.18013	4.0374	143	-0.24	019.00	0
019	Calcium (%)	2012	3.5250	0.03000	3.5686	0.18013	4.0374	143	-0.24	019.41	0
019	Calcium (%)	0675	3.5300	0.04000	3.5686	0.18013	4.0374	143	-0.21	019.43	0
019	Calcium (%)	0233	3.5300	0.16000	3.5686	0.18013	4.0374	143	-0.21	019.31	0
019	Calcium (%)	0940	3.5350	0.11000	3.5686	0.18013	4.0374	143	-0.19	019.08	0
019	Calcium (%)	0175	3.5350	0.01000	3.5686	0.18013	4.0374	143	-0.19	019.31	0
019	Calcium (%)	2006	3.5390	0.01600	3.5686	0.18013	4.0374	143	-0.16	019.00	0
019	Calcium (%)	0208	3.5410	0.00400	3.5686	0.18013	4.0374	143	-0.15	019.31	0
019	Calcium (%)	0098	3.5450	0.03000	3.5686	0.18013	4.0374	143	-0.13	019.41	0
019	Calcium (%)	0242	3.5450	0.01000	3.5686	0.18013	4.0374	143	-0.13	019.99	0
019	Calcium (%)	0138	3.5500	0.01400	3.5686	0.18013	4.0374	143	-0.10	019.08	0
019	Calcium (%)	0852	3.5500	0.10000	3.5686	0.18013	4.0374	143	-0.10	019.08	0
019	Calcium (%)	0300	3.5500	0.13200	3.5686	0.18013	4.0374	143	-0.10	019.41	0
019	Calcium (%)	0413	3.5500	0.10000	3.5686	0.18013	4.0374	143	-0.10	019.41	0
019	Calcium (%)	0910	3.5600	0.20000	3.5686	0.18013	4.0374	143	-0.05	019.41	0
019	Calcium (%)	0098	3.5650	0.11000	3.5686	0.18013	4.0374	143	-0.02	019.43	0
019	Calcium (%)	0689	3.5700	0.04000	3.5686	0.18013	4.0374	143	0.01	019.08	0
019	Calcium (%)	0045	3.5700	0.08000	3.5686	0.18013	4.0374	143	0.01	019.42	0
019	Calcium (%)	0169	3.5700	0.06000	3.5686	0.18013	4.0374	143	0.01	019.32	0
019	Calcium (%)	0038	3.5700	0.02000	3.5686	0.18013	4.0374	143	0.01	019.43	0
019	Calcium (%)	0186	3.5834	0.04400	3.5686	0.18013	4.0374	143	0.08	019.42	0
019	Calcium (%)	0536	3.5900	0.50000	3.5686	0.18013	4.0374	143	0.12	019.31	0
019	Calcium (%)	0689	3.5950	0.05000	3.5686	0.18013	4.0374	143	0.15	019.31	0
019	Calcium (%)	0948	3.5950	0.01000	3.5686	0.18013	4.0374	143	0.15	019.31	0
019	Calcium (%)	0598	3.5975	0.05300	3.5686	0.18013	4.0374	143	0.16	019.41	0
019	Calcium (%)	0674	3.6000	0.00000	3.5686	0.18013	4.0374	143	0.17	019.31	0
019	Calcium (%)	0693	3.6000	0.20000	3.5686	0.18013	4.0374	143	0.17	019.42	0
019	Calcium (%)	0148	3.6015	0.00300	3.5686	0.18013	4.0374	143	0.18	019.41	0
019	Calcium (%)	0918	3.6015	0.03700	3.5686	0.18013	4.0374	143	0.18	019.43	0
019	Calcium (%)	0893	3.6050	0.01000	3.5686	0.18013	4.0374	143	0.20	019.00	0
019	Calcium (%)	0425	3.6050	0.01000	3.5686	0.18013	4.0374	143	0.20	019.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	0885	3.6072	0.01380	3.5686	0.18013	4.0374	143	0.21	019.08	0
019	Calcium (%)	0572	3.6100	0.00000	3.5686	0.18013	4.0374	143	0.23	019.53	0
019	Calcium (%)	0278	3.6100	0.02000	3.5686	0.18013	4.0374	143	0.23	019.42	0
019	Calcium (%)	0563	3.6142	0.01520	3.5686	0.18013	4.0374	143	0.25	019.31	0
019	Calcium (%)	0036	3.6154	0.05790	3.5686	0.18013	4.0374	143	0.26	019.03	0
019	Calcium (%)	0874	3.6200	0.06000	3.5686	0.18013	4.0374	143	0.29	019.31	0
019	Calcium (%)	0861	3.6250	0.09000	3.5686	0.18013	4.0374	143	0.31	019.43	0
019	Calcium (%)	0028	3.6300	0.02000	3.5686	0.18013	4.0374	143	0.34	019.43	0
019	Calcium (%)	0964	3.6325	0.16390	3.5686	0.18013	4.0374	143	0.35	019.43	0
019	Calcium (%)	0074	3.6400	0.02000	3.5686	0.18013	4.0374	143	0.40	019.41	0
019	Calcium (%)	0037	3.6400	0.04000	3.5686	0.18013	4.0374	143	0.40	019.42	0
019	Calcium (%)	2009	3.6422	0.01570	3.5686	0.18013	4.0374	143	0.41	019.08	0
019	Calcium (%)	0878	3.6450	0.01000	3.5686	0.18013	4.0374	143	0.42	019.31	0
019	Calcium (%)	0229	3.6500	0.06000	3.5686	0.18013	4.0374	143	0.45	019.41	0
019	Calcium (%)	0108	3.6550	0.43000	3.5686	0.18013	4.0374	143	0.48	019.31	0
019	Calcium (%)	0610	3.6600	0.12000	3.5686	0.18013	4.0374	143	0.51	019.43	0
019	Calcium (%)	0520	3.6700	0.06000	3.5686	0.18013	4.0374	143	0.56	019.41	0
019	Calcium (%)	0504	3.6710	0.04600	3.5686	0.18013	4.0374	143	0.57	019.42	0
019	Calcium (%)	0968	3.6790	0.01600	3.5686	0.18013	4.0374	143	0.61	019.43	0
019	Calcium (%)	0066	3.6800	0.06000	3.5686	0.18013	4.0374	143	0.62	019.31	0
019	Calcium (%)	0049	3.6800	0.02000	3.5686	0.18013	4.0374	143	0.62	019.41	0
019	Calcium (%)	0613	3.6850	0.09000	3.5686	0.18013	4.0374	143	0.65	019.42	0
019	Calcium (%)	0505	3.6900	0.18000	3.5686	0.18013	4.0374	143	0.67	019.31	0
019	Calcium (%)	0682	3.6900	0.00000	3.5686	0.18013	4.0374	143	0.67	019.43	0
019	Calcium (%)	0083	3.6950	0.09000	3.5686	0.18013	4.0374	143	0.70	019.43	0
019	Calcium (%)	0043	3.7000	0.20000	3.5686	0.18013	4.0374	143	0.73	019.00	0
019	Calcium (%)	2152	3.7000	0.02000	3.5686	0.18013	4.0374	143	0.73	019.44	0
019	Calcium (%)	2089	3.7300	0.04000	3.5686	0.18013	4.0374	143	0.90	019.43	0
019	Calcium (%)	0676	3.7350	0.05000	3.5686	0.18013	4.0374	143	0.92	019.99	0
019	Calcium (%)	0504	3.7450	0.00600	3.5686	0.18013	4.0374	143	0.98	019.32	0
019	Calcium (%)	0612	3.7600	0.06000	3.5686	0.18013	4.0374	143	1.06	019.32	0
019	Calcium (%)	0366	3.7600	0.06000	3.5686	0.18013	4.0374	143	1.06	019.42	0
019	Calcium (%)	0034	3.7629	0.29830	3.5686	0.18013	4.0374	143	1.08	019.43	0
019	Calcium (%)	0027	3.7655	0.11300	3.5686	0.18013	4.0374	143	1.09	019.43	0
019	Calcium (%)	0629	3.7750	0.01000	3.5686	0.18013	4.0374	143	1.15	019.43	0
019	Calcium (%)	0033	3.8200	0.00000	3.5686	0.18013	4.0374	143	1.40	019.43	0
019	Calcium (%)	2053	3.8250	0.27000	3.5686	0.18013	4.0374	143	1.42	019.42	0
019	Calcium (%)	0353	3.8250	0.13000	3.5686	0.18013	4.0374	143	1.42	019.43	0
019	Calcium (%)	0692	3.8350	0.37000	3.5686	0.18013	4.0374	143	1.48	019.42	0
019	Calcium (%)	0009	3.8389	0.07100	3.5686	0.18013	4.0374	143	1.50	019.42	0
019	Calcium (%)	0407	3.8614	0.02890	3.5686	0.18013	4.0374	143	1.63	019.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	0504	3.8675	0.05700	3.5686	0.18013	4.0374	143	1.66	019.33	0
019	Calcium (%)	0620	3.8757	0.11560	3.5686	0.18013	4.0374	143	1.70	019.00	0
019	Calcium (%)	0010	3.9100	0.18000	3.5686	0.18013	4.0374	143	1.90	019.33	0
019	Calcium (%)	0226	3.9450	0.09000	3.5686	0.18013	4.0374	143	2.09	019.41	0
019	Calcium (%)	0100	3.9500	0.02000	3.5686	0.18013	4.0374	143	2.12	019.99	0
019	Calcium (%)	0619	3.9850	0.05000	3.5686	0.18013	4.0374	143	2.31	019.41	0
019	Calcium (%)	0190	3.9850	0.01000	3.5686	0.18013	4.0374	143	2.31	019.42	0
019	Calcium (%)	0726	4.0235	0.00500	3.5686	0.18013	4.0374	143	2.53	019.42	0
019	Calcium (%)	2128	4.0475	0.04900	3.5686	0.18013	4.0374	143	2.66	019.02	0
019	Calcium (%)	0152	4.5900	0.04000	3.5686	0.18013	4.0374	143	5.67	019.31	0
019	Calcium (%)	0018	4.6350	0.15000	3.5686	0.18013	4.0374	143	5.92	019.31	0
019	Calcium (%)	0536	5.5000	0.12000	3.5686	0.18013	4.0374	143	10.72	019.02	0
019	Calcium (%)	0202	32,071	565.96	3.5686	0.18013	4.0374	143	178019.87	019.43	0
021	Cobalt (mg / kg (ppm))	0619	0.96900	0.03600	1.6367	0.47160	0.14592	27	-1.42	021.41	0
021	Cobalt (mg / kg (ppm))	0910	1.0200	0.00000	1.6367	0.47160	0.14592	27	-1.31	021.52	0
021	Cobalt (mg / kg (ppm))	0345	1.0500	0.06000	1.6367	0.47160	0.14592	27	-1.24	021.43	0
021	Cobalt (mg / kg (ppm))	0918	1.0713	0.07690	1.6367	0.47160	0.14592	27	-1.20	021.53	0
021	Cobalt (mg / kg (ppm))	0096	1.1419	0.06490	1.6367	0.47160	0.14592	27	-1.05	021.52	0
021	Cobalt (mg / kg (ppm))	2053	1.1550	0.01000	1.6367	0.47160	0.14592	27	-1.02	021.42	0
021	Cobalt (mg / kg (ppm))	0675	1.2100	0.02000	1.6367	0.47160	0.14592	27	-0.90	021.43	0
021	Cobalt (mg / kg (ppm))	0693	1.2550	0.01000	1.6367	0.47160	0.14592	27	-0.81	021.42	0
021	Cobalt (mg / kg (ppm))	0045	1.4050	0.19000	1.6367	0.47160	0.14592	27	-0.49	021.42	0
021	Cobalt (mg / kg (ppm))	0560	1.5150	0.21000	1.6367	0.47160	0.14592	27	-0.26	021.42	0
021	Cobalt (mg / kg (ppm))	0689	1.5500	0.10000	1.6367	0.47160	0.14592	27	-0.18	021.31	0
021	Cobalt (mg / kg (ppm))	0171	1.6000	0.00000	1.6367	0.47160	0.14592	27	-0.08	021.41	0
021	Cobalt (mg / kg (ppm))	0098	1.6105	0.18700	1.6367	0.47160	0.14592	27	-0.06	021.53	0
021	Cobalt (mg / kg (ppm))	0964	1.7481	0.23510	1.6367	0.47160	0.14592	27	0.24	021.43	0
021	Cobalt (mg / kg (ppm))	0164	1.7500	0.10000	1.6367	0.47160	0.14592	27	0.24	021.31	0
021	Cobalt (mg / kg (ppm))	0610	1.7550	0.05000	1.6367	0.47160	0.14592	27	0.25	021.34	0
021	Cobalt (mg / kg (ppm))	0199	1.8000	0.08000	1.6367	0.47160	0.14592	27	0.35	021.53	0
021	Cobalt (mg / kg (ppm))	0208	1.8200	0.12000	1.6367	0.47160	0.14592	27	0.39	021.31	0
021	Cobalt (mg / kg (ppm))	0038	1.9000	0.20000	1.6367	0.47160	0.14592	27	0.56	021.43	0
021	Cobalt (mg / kg (ppm))	0043	1.9150	0.15000	1.6367	0.47160	0.14592	27	0.59	021.43	0
021	Cobalt (mg / kg (ppm))	0510	1.9400	0.08000	1.6367	0.47160	0.14592	27	0.64	021.43	0
021	Cobalt (mg / kg (ppm))	0366	1.9500	0.90000	1.6367	0.47160	0.14592	27	0.66	021.32	0
021	Cobalt (mg / kg (ppm))	0553	2.0850	0.11000	1.6367	0.47160	0.14592	27	0.95	021.53	0
021	Cobalt (mg / kg (ppm))	0021	2.2000	0.20000	1.6367	0.47160	0.14592	27	1.19	021.53	0
021	Cobalt (mg / kg (ppm))	0169	2.2100	0.42000	1.6367	0.47160	0.14592	27	1.22	021.43	0
021	Cobalt (mg / kg (ppm))	0590	2.2200	0.02000	1.6367	0.47160	0.14592	27	1.24	021.99	0
021	Cobalt (mg / kg (ppm))	0904	3.7150	0.31000	1.6367	0.47160	0.14592	27	4.41	021.31	0
021	Cobalt (mg / kg (ppm))	0021	1.1000	2.2000	1.6367	0.47160	0.14592	27		021.43	4

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
022	Copper (mg / kg (ppm))	0186	0.40000	0.40000	17.754	1.6809	1.1372	85	-10.32	022.42	0
022	Copper (mg / kg (ppm))	0405	1.8490	0.38200	17.754	1.6809	1.1372	85	-9.46	022.41	0
022	Copper (mg / kg (ppm))	2089	7.7400	3.3800	17.754	1.6809	1.1372	85	-5.96	022.43	0
022	Copper (mg / kg (ppm))	0353	8.5000	1.0000	17.754	1.6809	1.1372	85	-5.51	022.43	0
022	Copper (mg / kg (ppm))	0017	12.695	0.01000	17.754	1.6809	1.1372	85	-3.01	022.43	0
022	Copper (mg / kg (ppm))	0656	13.895	2.7700	17.754	1.6809	1.1372	85	-2.30	022.31	0
022	Copper (mg / kg (ppm))	0265	15.000	0.00000	17.754	1.6809	1.1372	85	-1.64	022.42	0
022	Copper (mg / kg (ppm))	0202	15.190	0.36000	17.754	1.6809	1.1372	85	-1.53	022.43	0
022	Copper (mg / kg (ppm))	0038	15.500	1.0000	17.754	1.6809	1.1372	85	-1.34	022.43	0
022	Copper (mg / kg (ppm))	2053	15.570	0.22000	17.754	1.6809	1.1372	85	-1.30	022.42	0
022	Copper (mg / kg (ppm))	0520	15.740	0.02000	17.754	1.6809	1.1372	85	-1.20	022.41	0
022	Copper (mg / kg (ppm))	0199	15.870	0.16000	17.754	1.6809	1.1372	85	-1.12	022.53	0
022	Copper (mg / kg (ppm))	0045	15.950	0.90000	17.754	1.6809	1.1372	85	-1.07	022.42	0
022	Copper (mg / kg (ppm))	0205	16.050	0.70000	17.754	1.6809	1.1372	85	-1.01	022.42	0
022	Copper (mg / kg (ppm))	0009	16.065	1.3230	17.754	1.6809	1.1372	85	-1.01	022.42	0
022	Copper (mg / kg (ppm))	0511	16.100	0.80000	17.754	1.6809	1.1372	85	-0.98	022.41	0
022	Copper (mg / kg (ppm))	0629	16.200	0.00000	17.754	1.6809	1.1372	85	-0.92	022.43	0
022	Copper (mg / kg (ppm))	0294	16.260	0.08000	17.754	1.6809	1.1372	85	-0.89	022.42	0
022	Copper (mg / kg (ppm))	0345	16.350	0.30000	17.754	1.6809	1.1372	85	-0.84	022.43	0
022	Copper (mg / kg (ppm))	0164	16.650	0.10000	17.754	1.6809	1.1372	85	-0.66	022.41	0
022	Copper (mg / kg (ppm))	0939	16.800	0.02000	17.754	1.6809	1.1372	85	-0.57	022.31	0
022	Copper (mg / kg (ppm))	0590	16.900	1.0000	17.754	1.6809	1.1372	85	-0.51	022.99	0
022	Copper (mg / kg (ppm))	0098	16.920	1.4400	17.754	1.6809	1.1372	85	-0.50	022.41	0
022	Copper (mg / kg (ppm))	0675	16.970	0.48000	17.754	1.6809	1.1372	85	-0.47	022.43	0
022	Copper (mg / kg (ppm))	0096	16.972	1.4558	17.754	1.6809	1.1372	85	-0.47	022.42	0
022	Copper (mg / kg (ppm))	0629	17.000	0.00000	17.754	1.6809	1.1372	85	-0.45	022.31	0
022	Copper (mg / kg (ppm))	0100	17.000	2.0000	17.754	1.6809	1.1372	85	-0.45	022.99	0
022	Copper (mg / kg (ppm))	0043	17.050	0.90000	17.754	1.6809	1.1372	85	-0.42	022.43	0
022	Copper (mg / kg (ppm))	0169	17.200	0.20000	17.754	1.6809	1.1372	85	-0.33	022.43	0
022	Copper (mg / kg (ppm))	0190	17.245	0.49000	17.754	1.6809	1.1372	85	-0.30	022.42	0
022	Copper (mg / kg (ppm))	0682	17.300	0.00000	17.754	1.6809	1.1372	85	-0.27	022.43	0
022	Copper (mg / kg (ppm))	0098	17.325	0.47000	17.754	1.6809	1.1372	85	-0.26	022.43	0
022	Copper (mg / kg (ppm))	0619	17.400	1.4000	17.754	1.6809	1.1372	85	-0.21	022.41	0
022	Copper (mg / kg (ppm))	0425	17.400	0.20000	17.754	1.6809	1.1372	85	-0.21	022.43	0
022	Copper (mg / kg (ppm))	2149	17.485	0.43000	17.754	1.6809	1.1372	85	-0.16	022.31	0
022	Copper (mg / kg (ppm))	0175	17.500	1.0000	17.754	1.6809	1.1372	85	-0.15	022.31	0
022	Copper (mg / kg (ppm))	0083	17.500	1.0000	17.754	1.6809	1.1372	85	-0.15	022.43	0
022	Copper (mg / kg (ppm))	0510	17.500	1.0000	17.754	1.6809	1.1372	85	-0.15	022.43	0
022	Copper (mg / kg (ppm))	0613	17.610	2.2800	17.754	1.6809	1.1372	85	-0.09	022.42	0
022	Copper (mg / kg (ppm))	0726	17.685	0.07000	17.754	1.6809	1.1372	85	-0.04	022.42	0
022	Copper (mg / kg (ppm))	0278	17.750	0.10000	17.754	1.6809	1.1372	85	0.00	022.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
022	Copper (mg / kg (ppm))	0413	17.800	0.20000	17.754	1.6809	1.1372	85	0.03	022.42	0
022	Copper (mg / kg (ppm))	0357	17.834	1.0660	17.754	1.6809	1.1372	85	0.05	022.42	0
022	Copper (mg / kg (ppm))	0096	17.843	0.48580	17.754	1.6809	1.1372	85	0.05	022.52	0
022	Copper (mg / kg (ppm))	0918	17.980	1.9400	17.754	1.6809	1.1372	85	0.13	022.43	0
022	Copper (mg / kg (ppm))	0019	18.000	3.0000	17.754	1.6809	1.1372	85	0.15	022.41	0
022	Copper (mg / kg (ppm))	0610	18.000	0.00000	17.754	1.6809	1.1372	85	0.15	022.43	0
022	Copper (mg / kg (ppm))	0968	18.000	0.00000	17.754	1.6809	1.1372	85	0.15	022.43	0
022	Copper (mg / kg (ppm))	0358	18.030	1.7600	17.754	1.6809	1.1372	85	0.16	022.41	0
022	Copper (mg / kg (ppm))	0693	18.050	0.50000	17.754	1.6809	1.1372	85	0.18	022.42	0
022	Copper (mg / kg (ppm))	0553	18.100	2.2000	17.754	1.6809	1.1372	85	0.21	022.53	0
022	Copper (mg / kg (ppm))	0148	18.125	0.29000	17.754	1.6809	1.1372	85	0.22	022.41	0
022	Copper (mg / kg (ppm))	0572	18.150	0.50000	17.754	1.6809	1.1372	85	0.24	022.53	0
022	Copper (mg / kg (ppm))	0407	18.198	0.24340	17.754	1.6809	1.1372	85	0.26	022.41	0
022	Copper (mg / kg (ppm))	0904	18.300	0.48000	17.754	1.6809	1.1372	85	0.32	022.31	0
022	Copper (mg / kg (ppm))	0563	18.301	0.53600	17.754	1.6809	1.1372	85	0.33	022.31	0
022	Copper (mg / kg (ppm))	0229	18.475	2.0500	17.754	1.6809	1.1372	85	0.43	022.41	0
022	Copper (mg / kg (ppm))	0074	18.500	1.0000	17.754	1.6809	1.1372	85	0.44	022.41	0
022	Copper (mg / kg (ppm))	0226	18.500	1.0000	17.754	1.6809	1.1372	85	0.44	022.41	0
022	Copper (mg / kg (ppm))	0021	18.500	3.0000	17.754	1.6809	1.1372	85	0.44	022.53	0
022	Copper (mg / kg (ppm))	0208	18.535	3.5300	17.754	1.6809	1.1372	85	0.46	022.41	0
022	Copper (mg / kg (ppm))	0066	18.580	0.76000	17.754	1.6809	1.1372	85	0.49	022.31	0
022	Copper (mg / kg (ppm))	0692	18.700	1.0000	17.754	1.6809	1.1372	85	0.56	022.42	0
022	Copper (mg / kg (ppm))	0861	18.850	0.30000	17.754	1.6809	1.1372	85	0.65	022.43	0
022	Copper (mg / kg (ppm))	0878	18.900	0.20000	17.754	1.6809	1.1372	85	0.68	022.31	0
022	Copper (mg / kg (ppm))	0171	18.900	1.0000	17.754	1.6809	1.1372	85	0.68	022.41	0
022	Copper (mg / kg (ppm))	0560	18.900	2.6000	17.754	1.6809	1.1372	85	0.68	022.42	0
022	Copper (mg / kg (ppm))	0910	19.000	4.0000	17.754	1.6809	1.1372	85	0.74	022.41	0
022	Copper (mg / kg (ppm))	0366	19.000	0.00000	17.754	1.6809	1.1372	85	0.74	022.42	0
022	Copper (mg / kg (ppm))	0948	19.255	0.37000	17.754	1.6809	1.1372	85	0.89	022.33	0
022	Copper (mg / kg (ppm))	0049	19.400	1.4400	17.754	1.6809	1.1372	85	0.98	022.41	0
022	Copper (mg / kg (ppm))	0039	19.464	6.3650	17.754	1.6809	1.1372	85	1.02	022.44	0
022	Copper (mg / kg (ppm))	0242	19.500	3.0000	17.754	1.6809	1.1372	85	1.04	022.99	0
022	Copper (mg / kg (ppm))	0187	19.635	0.27000	17.754	1.6809	1.1372	85	1.12	022.42	0
022	Copper (mg / kg (ppm))	2012	19.675	0.67000	17.754	1.6809	1.1372	85	1.14	022.41	0
022	Copper (mg / kg (ppm))	0598	19.735	0.29000	17.754	1.6809	1.1372	85	1.18	022.41	0
022	Copper (mg / kg (ppm))	0964	19.777	2.6058	17.754	1.6809	1.1372	85	1.20	022.43	0
022	Copper (mg / kg (ppm))	0529	20.000	1.5000	17.754	1.6809	1.1372	85	1.34	022.31	0
022	Copper (mg / kg (ppm))	0037	20.650	0.30000	17.754	1.6809	1.1372	85	1.72	022.42	0
022	Copper (mg / kg (ppm))	0038	21.000	0.00000	17.754	1.6809	1.1372	85	1.93	022.32	0
022	Copper (mg / kg (ppm))	0504	21.105	5.7700	17.754	1.6809	1.1372	85	1.99	022.32	0
022	Copper (mg / kg (ppm))	0354	21.190	0.80000	17.754	1.6809	1.1372	85	2.04	022.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
022	Copper (mg / kg (ppm))	2062	21.650	7.5000	17.754	1.6809	1.1372	85	2.32	022.31	0
022	Copper (mg / kg (ppm))	0884	21.700	0.00000	17.754	1.6809	1.1372	85	2.35	022.31	0
022	Copper (mg / kg (ppm))	0689	23.850	2.3000	17.754	1.6809	1.1372	85	3.63	022.31	0
022	Copper (mg / kg (ppm))	0021	0.00000	0.00000	17.754	1.6809	1.1372	85		022.43	4
023	Fluorine (mg / kg (ppm))	0910	8.1500	0.30000				1		023.01	0
025	Iron (mg / kg (ppm))	0405	21.565	0.05000	312.85	40.109	17.157	84	-7.26	025.41	0
025	Iron (mg / kg (ppm))	0187	82.790	2.4200	312.85	40.109	17.157	84	-5.74	025.42	0
025	Iron (mg / kg (ppm))	0726	211.00	5.6000	312.85	40.109	17.157	84	-2.54	025.42	0
025	Iron (mg / kg (ppm))	2053	234.50	3.0000	312.85	40.109	17.157	84	-1.95	025.42	0
025	Iron (mg / kg (ppm))	0560	238.55	2.1000	312.85	40.109	17.157	84	-1.85	025.42	0
025	Iron (mg / kg (ppm))	0968	241.50	3.0000	312.85	40.109	17.157	84	-1.78	025.43	0
025	Iron (mg / kg (ppm))	2089	245.39	33.700	312.85	40.109	17.157	84	-1.68	025.43	0
025	Iron (mg / kg (ppm))	0190	246.30	9.2000	312.85	40.109	17.157	84	-1.66	025.42	0
025	Iron (mg / kg (ppm))	0154	252.91	16.890	312.85	40.109	17.157	84	-1.49	025.52	0
025	Iron (mg / kg (ppm))	2012	255.85	5.5500	312.85	40.109	17.157	84	-1.42	025.41	0
025	Iron (mg / kg (ppm))	0199	257.95	6.3000	312.85	40.109	17.157	84	-1.37	025.53	0
025	Iron (mg / kg (ppm))	0043	264.00	8.0000	312.85	40.109	17.157	84	-1.22	025.43	0
025	Iron (mg / kg (ppm))	0096	268.38	73.523	312.85	40.109	17.157	84	-1.11	025.42	0
025	Iron (mg / kg (ppm))	0345	269.50	9.0000	312.85	40.109	17.157	84	-1.08	025.43	0
025	Iron (mg / kg (ppm))	0619	270.50	7.0000	312.85	40.109	17.157	84	-1.06	025.41	0
025	Iron (mg / kg (ppm))	0265	270.50	19.000	312.85	40.109	17.157	84	-1.06	025.42	0
025	Iron (mg / kg (ppm))	0598	277.05	6.9000	312.85	40.109	17.157	84	-0.89	025.41	0
025	Iron (mg / kg (ppm))	0017	277.90	46.910	312.85	40.109	17.157	84	-0.87	025.43	0
025	Iron (mg / kg (ppm))	0353	278.00	6.0000	312.85	40.109	17.157	84	-0.87	025.43	0
025	Iron (mg / kg (ppm))	0294	280.01	12.820	312.85	40.109	17.157	84	-0.82	025.42	0
025	Iron (mg / kg (ppm))	0629	284.00	4.0000	312.85	40.109	17.157	84	-0.72	025.43	0
025	Iron (mg / kg (ppm))	0510	285.50	5.0000	312.85	40.109	17.157	84	-0.68	025.43	0
025	Iron (mg / kg (ppm))	0629	286.00	0.00000	312.85	40.109	17.157	84	-0.67	025.31	0
025	Iron (mg / kg (ppm))	0693	290.00	10.000	312.85	40.109	17.157	84	-0.57	025.42	0
025	Iron (mg / kg (ppm))	0413	294.50	7.0000	312.85	40.109	17.157	84	-0.46	025.42	0
025	Iron (mg / kg (ppm))	0186	294.85	29.700	312.85	40.109	17.157	84	-0.45	025.42	0
025	Iron (mg / kg (ppm))	0692	295.00	18.000	312.85	40.109	17.157	84	-0.45	025.42	0
025	Iron (mg / kg (ppm))	0689	299.35	9.7000	312.85	40.109	17.157	84	-0.34	025.31	0
025	Iron (mg / kg (ppm))	0083	300.00	6.0000	312.85	40.109	17.157	84	-0.32	025.43	0
025	Iron (mg / kg (ppm))	0039	300.12	16.104	312.85	40.109	17.157	84	-0.32	025.41	0
025	Iron (mg / kg (ppm))	0100	304.50	29.000	312.85	40.109	17.157	84	-0.21	025.99	0
025	Iron (mg / kg (ppm))	0675	304.90	8.2700	312.85	40.109	17.157	84	-0.20	025.43	0
025	Iron (mg / kg (ppm))	0948	306.87	4.3340	312.85	40.109	17.157	84	-0.15	025.33	0
025	Iron (mg / kg (ppm))	0590	307.00	8.0000	312.85	40.109	17.157	84	-0.15	025.99	0
025	Iron (mg / kg (ppm))	0610	308.00	0.00000	312.85	40.109	17.157	84	-0.12	025.43	0
025	Iron (mg / kg (ppm))	0045	313.00	8.0000	312.85	40.109	17.157	84	0.00	025.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
025	Iron (mg / kg (ppm))	0242	313.00	40.000	312.85	40.109	17.157	84	0.00	025.99	0
025	Iron (mg / kg (ppm))	0169	316.00	8.0000	312.85	40.109	17.157	84	0.08	025.43	0
025	Iron (mg / kg (ppm))	0407	316.43	8.0272	312.85	40.109	17.157	84	0.09	025.41	0
025	Iron (mg / kg (ppm))	0878	318.00	0.00000	312.85	40.109	17.157	84	0.13	025.31	0
025	Iron (mg / kg (ppm))	0520	318.00	56.200	312.85	40.109	17.157	84	0.13	025.41	0
025	Iron (mg / kg (ppm))	0504	318.30	11.400	312.85	40.109	17.157	84	0.14	025.33	0
025	Iron (mg / kg (ppm))	0563	318.50	17.000	312.85	40.109	17.157	84	0.14	025.31	0
025	Iron (mg / kg (ppm))	0613	321.75	49.700	312.85	40.109	17.157	84	0.22	025.42	0
025	Iron (mg / kg (ppm))	0098	324.25	14.300	312.85	40.109	17.157	84	0.28	025.43	0
025	Iron (mg / kg (ppm))	0511	325.00	20.000	312.85	40.109	17.157	84	0.30	025.41	0
025	Iron (mg / kg (ppm))	0226	325.50	13.000	312.85	40.109	17.157	84	0.32	025.41	0
025	Iron (mg / kg (ppm))	0910	326.00	16.000	312.85	40.109	17.157	84	0.33	025.41	0
025	Iron (mg / kg (ppm))	0964	327.09	54.958	312.85	40.109	17.157	84	0.36	025.43	0
025	Iron (mg / kg (ppm))	0037	327.20	11.600	312.85	40.109	17.157	84	0.36	025.42	0
025	Iron (mg / kg (ppm))	0171	328.50	15.000	312.85	40.109	17.157	84	0.39	025.41	0
025	Iron (mg / kg (ppm))	0148	328.85	1.5000	312.85	40.109	17.157	84	0.40	025.41	0
025	Iron (mg / kg (ppm))	0904	329.68	20.280	312.85	40.109	17.157	84	0.42	025.31	0
025	Iron (mg / kg (ppm))	0682	330.00	0.00000	312.85	40.109	17.157	84	0.43	025.43	0
025	Iron (mg / kg (ppm))	0425	330.15	7.5000	312.85	40.109	17.157	84	0.43	025.43	0
025	Iron (mg / kg (ppm))	0354	330.76	0.56000	312.85	40.109	17.157	84	0.45	025.31	0
025	Iron (mg / kg (ppm))	0038	333.00	0.00000	312.85	40.109	17.157	84	0.50	025.43	0
025	Iron (mg / kg (ppm))	0019	333.95	9.9000	312.85	40.109	17.157	84	0.53	025.41	0
025	Iron (mg / kg (ppm))	0164	334.50	3.0000	312.85	40.109	17.157	84	0.54	025.41	0
025	Iron (mg / kg (ppm))	0874	335.00	14.000	312.85	40.109	17.157	84	0.55	025.31	0
025	Iron (mg / kg (ppm))	0278	335.00	14.000	312.85	40.109	17.157	84	0.55	025.42	0
025	Iron (mg / kg (ppm))	0656	337.92	5.9300	312.85	40.109	17.157	84	0.62	025.31	0
025	Iron (mg / kg (ppm))	0208	338.00	42.000	312.85	40.109	17.157	84	0.63	025.31	0
025	Iron (mg / kg (ppm))	0358	338.45	27.340	312.85	40.109	17.157	84	0.64	025.41	0
025	Iron (mg / kg (ppm))	0038	339.50	3.0000	312.85	40.109	17.157	84	0.66	025.31	0
025	Iron (mg / kg (ppm))	0098	339.60	36.200	312.85	40.109	17.157	84	0.67	025.41	0
025	Iron (mg / kg (ppm))	2022	340.00	10.000	312.85	40.109	17.157	84	0.68	025.31	0
025	Iron (mg / kg (ppm))	0366	342.00	86.000	312.85	40.109	17.157	84	0.73	025.42	0
025	Iron (mg / kg (ppm))	0504	344.50	9.8000	312.85	40.109	17.157	84	0.79	025.32	0
025	Iron (mg / kg (ppm))	2062	344.90	2.0000	312.85	40.109	17.157	84	0.80	025.31	0
025	Iron (mg / kg (ppm))	0536	347.56	13.080	312.85	40.109	17.157	84	0.87	025.31	0
025	Iron (mg / kg (ppm))	0529	348.32	70.630	312.85	40.109	17.157	84	0.88	025.31	0
025	Iron (mg / kg (ppm))	0021	350.00	20.000	312.85	40.109	17.157	84	0.93	025.43	0
025	Iron (mg / kg (ppm))	0918	352.32	68.960	312.85	40.109	17.157	84	0.98	025.43	0
025	Iron (mg / kg (ppm))	0049	355.42	15.830	312.85	40.109	17.157	84	1.06	025.41	0
025	Iron (mg / kg (ppm))	0074	355.50	21.000	312.85	40.109	17.157	84	1.06	025.41	0
025	Iron (mg / kg (ppm))	0096	356.20	18.446	312.85	40.109	17.157	84	1.08	025.52	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
025	Iron (mg / kg (ppm))	0553	358.50	7.0000	312.85	40.109	17.157	84	1.14	025.53	0
025	Iron (mg / kg (ppm))	0066	364.49	11.860	312.85	40.109	17.157	84	1.29	025.31	0
025	Iron (mg / kg (ppm))	0229	372.18	11.100	312.85	40.109	17.157	84	1.48	025.41	0
025	Iron (mg / kg (ppm))	0208	376.80	19.000	312.85	40.109	17.157	84	1.59	025.41	0
025	Iron (mg / kg (ppm))	0175	380.00	10.000	312.85	40.109	17.157	84	1.67	025.31	0
025	Iron (mg / kg (ppm))	0861	389.00	80.000	312.85	40.109	17.157	84	1.90	025.43	0
025	Iron (mg / kg (ppm))	0884	445.60	6.0000	312.85	40.109	17.157	84	3.31	025.31	0
027	Magnesium (%)	0405	0.03990	0.00220	0.31103	0.01758	0.10302	101	-15.42	027.41	0
027	Magnesium (%)	0026	0.19445	0.00370	0.31103	0.01758	0.10302	101	-6.63	027.42	0
027	Magnesium (%)	0001	0.26765	0.00610	0.31103	0.01758	0.10302	101	-2.47	027.31	0
027	Magnesium (%)	0096	0.27295	0.00770	0.31103	0.01758	0.10302	101	-2.17	027.42	0
027	Magnesium (%)	0650	0.27330	0.03100	0.31103	0.01758	0.10302	101	-2.15	027.31	0
027	Magnesium (%)	0014	0.27650	0.00500	0.31103	0.01758	0.10302	101	-1.96	027.31	0
027	Magnesium (%)	2129	0.28350	0.00100	0.31103	0.01758	0.10302	101	-1.57	027.42	0
027	Magnesium (%)	0021	0.28500	0.03000	0.31103	0.01758	0.10302	101	-1.48	027.43	0
027	Magnesium (%)	0675	0.28500	0.01000	0.31103	0.01758	0.10302	101	-1.48	027.43	0
027	Magnesium (%)	0038	0.28650	0.00900	0.31103	0.01758	0.10302	101	-1.40	027.43	0
027	Magnesium (%)	0354	0.29000	0.00000	0.31103	0.01758	0.10302	101	-1.20	027.31	0
027	Magnesium (%)	0520	0.29150	0.00500	0.31103	0.01758	0.10302	101	-1.11	027.41	0
027	Magnesium (%)	0154	0.29245	0.01630	0.31103	0.01758	0.10302	101	-1.06	027.52	0
027	Magnesium (%)	0039	0.29400	0.01400	0.31103	0.01758	0.10302	101	-0.97	027.44	0
027	Magnesium (%)	0169	0.29450	0.00300	0.31103	0.01758	0.10302	101	-0.94	027.32	0
027	Magnesium (%)	0345	0.29450	0.00700	0.31103	0.01758	0.10302	101	-0.94	027.43	0
027	Magnesium (%)	0263	0.29460	0.00000	0.31103	0.01758	0.10302	101	-0.93	027.32	0
027	Magnesium (%)	0536	0.29500	0.01000	0.31103	0.01758	0.10302	101	-0.91	027.31	0
027	Magnesium (%)	0294	0.29500	0.01000	0.31103	0.01758	0.10302	101	-0.91	027.42	0
027	Magnesium (%)	0038	0.29600	0.00000	0.31103	0.01758	0.10302	101	-0.85	027.31	0
027	Magnesium (%)	0504	0.29650	0.00700	0.31103	0.01758	0.10302	101	-0.83	027.33	0
027	Magnesium (%)	0142	0.29820	0.00080	0.31103	0.01758	0.10302	101	-0.73	027.31	0
027	Magnesium (%)	0148	0.29900	0.00200	0.31103	0.01758	0.10302	101	-0.68	027.41	0
027	Magnesium (%)	0910	0.30000	0.02000	0.31103	0.01758	0.10302	101	-0.63	027.41	0
027	Magnesium (%)	2012	0.30000	0.00000	0.31103	0.01758	0.10302	101	-0.63	027.41	0
027	Magnesium (%)	0510	0.30000	0.00000	0.31103	0.01758	0.10302	101	-0.63	027.43	0
027	Magnesium (%)	0242	0.30000	0.00000	0.31103	0.01758	0.10302	101	-0.63	027.99	0
027	Magnesium (%)	0009	0.30155	0.01090	0.31103	0.01758	0.10302	101	-0.54	027.42	0
027	Magnesium (%)	0357	0.30175	0.00830	0.31103	0.01758	0.10302	101	-0.53	027.42	0
027	Magnesium (%)	2022	0.30200	0.00000	0.31103	0.01758	0.10302	101	-0.51	027.31	0
027	Magnesium (%)	0164	0.30250	0.00300	0.31103	0.01758	0.10302	101	-0.49	027.41	0
027	Magnesium (%)	0560	0.30395	0.00370	0.31103	0.01758	0.10302	101	-0.40	027.42	0
027	Magnesium (%)	0066	0.30500	0.01000	0.31103	0.01758	0.10302	101	-0.34	027.31	0
027	Magnesium (%)	0175	0.30500	0.01000	0.31103	0.01758	0.10302	101	-0.34	027.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
027	Magnesium (%)	0904	0.30500	0.01000	0.31103	0.01758	0.10302	101	-0.34	027.31	0
027	Magnesium (%)	0049	0.30500	0.01000	0.31103	0.01758	0.10302	101	-0.34	027.41	0
027	Magnesium (%)	0098	0.30500	0.01000	0.31103	0.01758	0.10302	101	-0.34	027.41	0
027	Magnesium (%)	0413	0.30500	0.01000	0.31103	0.01758	0.10302	101	-0.34	027.41	0
027	Magnesium (%)	0613	0.30500	0.01000	0.31103	0.01758	0.10302	101	-0.34	027.42	0
027	Magnesium (%)	0035	0.30500	0.01000	0.31103	0.01758	0.10302	101	-0.34	027.43	0
027	Magnesium (%)	0353	0.30500	0.01000	0.31103	0.01758	0.10302	101	-0.34	027.43	0
027	Magnesium (%)	0889	0.30500	0.01000	0.31103	0.01758	0.10302	101	-0.34	027.99	0
027	Magnesium (%)	0874	0.30550	0.00500	0.31103	0.01758	0.10302	101	-0.31	027.31	0
027	Magnesium (%)	0504	0.30800	0.00800	0.31103	0.01758	0.10302	101	-0.17	027.32	0
027	Magnesium (%)	0043	0.30800	0.00200	0.31103	0.01758	0.10302	101	-0.17	027.43	0
027	Magnesium (%)	0598	0.30860	0.00200	0.31103	0.01758	0.10302	101	-0.14	027.41	0
027	Magnesium (%)	0563	0.30910	0.00080	0.31103	0.01758	0.10302	101	-0.11	027.31	0
027	Magnesium (%)	0208	0.31000	0.01200	0.31103	0.01758	0.10302	101	-0.06	027.31	0
027	Magnesium (%)	0629	0.31000	0.00000	0.31103	0.01758	0.10302	101	-0.06	027.31	0
027	Magnesium (%)	0265	0.31000	0.00000	0.31103	0.01758	0.10302	101	-0.06	027.42	0
027	Magnesium (%)	0366	0.31000	0.00000	0.31103	0.01758	0.10302	101	-0.06	027.42	0
027	Magnesium (%)	0028	0.31000	0.00000	0.31103	0.01758	0.10302	101	-0.06	027.43	0
027	Magnesium (%)	0098	0.31000	0.00000	0.31103	0.01758	0.10302	101	-0.06	027.43	0
027	Magnesium (%)	0425	0.31000	0.00000	0.31103	0.01758	0.10302	101	-0.06	027.43	0
027	Magnesium (%)	0682	0.31000	0.00000	0.31103	0.01758	0.10302	101	-0.06	027.43	0
027	Magnesium (%)	0590	0.31000	0.00000	0.31103	0.01758	0.10302	101	-0.06	027.99	0
027	Magnesium (%)	0171	0.31100	0.00600	0.31103	0.01758	0.10302	101	0.00	027.41	0
027	Magnesium (%)	0918	0.31250	0.01100	0.31103	0.01758	0.10302	101	0.08	027.43	0
027	Magnesium (%)	0139	0.31380	0.00100	0.31103	0.01758	0.10302	101	0.16	027.41	0
027	Magnesium (%)	0096	0.31415	0.00930	0.31103	0.01758	0.10302	101	0.18	027.52	0
027	Magnesium (%)	0529	0.31500	0.05000	0.31103	0.01758	0.10302	101	0.23	027.31	0
027	Magnesium (%)	0019	0.31500	0.01000	0.31103	0.01758	0.10302	101	0.23	027.41	0
027	Magnesium (%)	0358	0.31500	0.01000	0.31103	0.01758	0.10302	101	0.23	027.41	0
027	Magnesium (%)	0511	0.31500	0.01000	0.31103	0.01758	0.10302	101	0.23	027.41	0
027	Magnesium (%)	0278	0.31500	0.01000	0.31103	0.01758	0.10302	101	0.23	027.42	0
027	Magnesium (%)	0033	0.31500	0.01000	0.31103	0.01758	0.10302	101	0.23	027.43	0
027	Magnesium (%)	0693	0.31900	0.00000	0.31103	0.01758	0.10302	101	0.45	027.42	0
027	Magnesium (%)	0884	0.32000	0.00000	0.31103	0.01758	0.10302	101	0.51	027.31	0
027	Magnesium (%)	0229	0.32000	0.00000	0.31103	0.01758	0.10302	101	0.51	027.41	0
027	Magnesium (%)	0610	0.32000	0.00000	0.31103	0.01758	0.10302	101	0.51	027.43	0
027	Magnesium (%)	0629	0.32000	0.00000	0.31103	0.01758	0.10302	101	0.51	027.43	0
027	Magnesium (%)	2089	0.32000	0.00000	0.31103	0.01758	0.10302	101	0.51	027.43	0
027	Magnesium (%)	0045	0.32100	0.00000	0.31103	0.01758	0.10302	101	0.57	027.42	0
027	Magnesium (%)	0074	0.32200	0.00200	0.31103	0.01758	0.10302	101	0.62	027.41	0
027	Magnesium (%)	0187	0.32220	0.00300	0.31103	0.01758	0.10302	101	0.64	027.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
027	Magnesium (%)	0027	0.32300	0.00200	0.31103	0.01758	0.10302	101	0.68	027.43	0
027	Magnesium (%)	0726	0.32375	0.00030	0.31103	0.01758	0.10302	101	0.72	027.42	0
027	Magnesium (%)	0035	0.32500	0.01000	0.31103	0.01758	0.10302	101	0.80	027.33	0
027	Magnesium (%)	0948	0.32500	0.01000	0.31103	0.01758	0.10302	101	0.80	027.33	0
027	Magnesium (%)	0037	0.32500	0.01000	0.31103	0.01758	0.10302	101	0.80	027.42	0
027	Magnesium (%)	0190	0.32500	0.01000	0.31103	0.01758	0.10302	101	0.80	027.42	0
027	Magnesium (%)	0553	0.32650	0.00900	0.31103	0.01758	0.10302	101	0.88	027.53	0
027	Magnesium (%)	0572	0.32650	0.01700	0.31103	0.01758	0.10302	101	0.88	027.53	0
027	Magnesium (%)	0407	0.32855	0.00510	0.31103	0.01758	0.10302	101	1.00	027.41	0
027	Magnesium (%)	0692	0.33000	0.04000	0.31103	0.01758	0.10302	101	1.08	027.42	0
027	Magnesium (%)	0878	0.33000	0.00000	0.31103	0.01758	0.10302	101	1.08	027.31	0
027	Magnesium (%)	0619	0.33100	0.00200	0.31103	0.01758	0.10302	101	1.14	027.41	0
027	Magnesium (%)	0861	0.33150	0.00500	0.31103	0.01758	0.10302	101	1.16	027.43	0
027	Magnesium (%)	0186	0.33215	0.03090	0.31103	0.01758	0.10302	101	1.20	027.42	0
027	Magnesium (%)	0208	0.33250	0.01500	0.31103	0.01758	0.10302	101	1.22	027.41	0
027	Magnesium (%)	0034	0.33445	0.01690	0.31103	0.01758	0.10302	101	1.33	027.43	0
027	Magnesium (%)	0505	0.33500	0.00200	0.31103	0.01758	0.10302	101	1.36	027.31	0
027	Magnesium (%)	2053	0.33500	0.01000	0.31103	0.01758	0.10302	101	1.36	027.42	0
027	Magnesium (%)	0083	0.33500	0.01000	0.31103	0.01758	0.10302	101	1.36	027.43	0
027	Magnesium (%)	0964	0.34090	0.00860	0.31103	0.01758	0.10302	101	1.70	027.43	0
027	Magnesium (%)	0689	0.34500	0.01000	0.31103	0.01758	0.10302	101	1.93	027.31	0
027	Magnesium (%)	0226	0.35000	0.00000	0.31103	0.01758	0.10302	101	2.22	027.41	0
027	Magnesium (%)	0100	0.36500	0.01000	0.31103	0.01758	0.10302	101	3.07	027.99	0
027	Magnesium (%)	0968	0.38500	0.00200	0.31103	0.01758	0.10302	101	4.21	027.43	0
027	Magnesium (%)	0017	149.22	3.3100	0.31103	0.01758	0.10302	101	8471.34	027.43	0
027	Magnesium (%)	0202	2,649.9	6.3700	0.31103	0.01758	0.10302	101	150737.65	027.43	0
028	Manganese (mg / kg (ppm))	0405	11.170	0.50000	162.84	10.938	5.4585	95	-13.87	028.41	0
028	Manganese (mg / kg (ppm))	2062	16.550	0.30000	162.84	10.938	5.4585	95	-13.37	028.31	0
028	Manganese (mg / kg (ppm))	0026	78.500	3.0000	162.84	10.938	5.4585	95	-7.71	028.42	0
028	Manganese (mg / kg (ppm))	0619	123.50	1.0000	162.84	10.938	5.4585	95	-3.60	028.41	0
028	Manganese (mg / kg (ppm))	2053	136.00	0.00000	162.84	10.938	5.4585	95	-2.45	028.42	0
028	Manganese (mg / kg (ppm))	0009	138.81	2.6270	162.84	10.938	5.4585	95	-2.20	028.42	0
028	Manganese (mg / kg (ppm))	0598	139.50	1.6000	162.84	10.938	5.4585	95	-2.13	028.41	0
028	Manganese (mg / kg (ppm))	2022	144.00	0.00000	162.84	10.938	5.4585	95	-1.72	028.31	0
028	Manganese (mg / kg (ppm))	0689	144.05	9.9000	162.84	10.938	5.4585	95	-1.72	028.31	0
028	Manganese (mg / kg (ppm))	0202	144.71	4.7600	162.84	10.938	5.4585	95	-1.66	028.43	0
028	Manganese (mg / kg (ppm))	0226	145.00	18.000	162.84	10.938	5.4585	95	-1.63	028.41	0
028	Manganese (mg / kg (ppm))	0613	148.70	8.4000	162.84	10.938	5.4585	95	-1.29	028.42	0
028	Manganese (mg / kg (ppm))	0353	149.50	9.0000	162.84	10.938	5.4585	95	-1.22	028.43	0
028	Manganese (mg / kg (ppm))	0948	149.93	0.52100	162.84	10.938	5.4585	95	-1.18	028.33	0
028	Manganese (mg / kg (ppm))	0096	150.89	14.054	162.84	10.938	5.4585	95	-1.09	028.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
028	Manganese (mg / kg (ppm))	0043	152.50	9.0000	162.84	10.938	5.4585	95	-0.95	028.43	0
028	Manganese (mg / kg (ppm))	0693	153.50	15.000	162.84	10.938	5.4585	95	-0.85	028.42	0
028	Manganese (mg / kg (ppm))	2012	154.21	5.3200	162.84	10.938	5.4585	95	-0.79	028.41	0
028	Manganese (mg / kg (ppm))	0520	155.55	7.7000	162.84	10.938	5.4585	95	-0.67	028.41	0
028	Manganese (mg / kg (ppm))	0171	156.00	2.0000	162.84	10.938	5.4585	95	-0.63	028.41	0
028	Manganese (mg / kg (ppm))	0510	156.00	2.0000	162.84	10.938	5.4585	95	-0.63	028.43	0
028	Manganese (mg / kg (ppm))	0164	156.50	1.0000	162.84	10.938	5.4585	95	-0.58	028.41	0
028	Manganese (mg / kg (ppm))	0278	156.50	3.0000	162.84	10.938	5.4585	95	-0.58	028.42	0
028	Manganese (mg / kg (ppm))	0049	156.84	9.4800	162.84	10.938	5.4585	95	-0.55	028.41	0
028	Manganese (mg / kg (ppm))	0014	157.00	6.0000	162.84	10.938	5.4585	95	-0.53	028.31	0
028	Manganese (mg / kg (ppm))	0511	157.00	4.0000	162.84	10.938	5.4585	95	-0.53	028.41	0
028	Manganese (mg / kg (ppm))	0345	157.00	4.0000	162.84	10.938	5.4585	95	-0.53	028.43	0
028	Manganese (mg / kg (ppm))	0968	157.00	2.0000	162.84	10.938	5.4585	95	-0.53	028.43	0
028	Manganese (mg / kg (ppm))	0019	157.50	3.0000	162.84	10.938	5.4585	95	-0.49	028.41	0
028	Manganese (mg / kg (ppm))	0148	157.70	1.8000	162.84	10.938	5.4585	95	-0.47	028.41	0
028	Manganese (mg / kg (ppm))	0413	158.00	2.0000	162.84	10.938	5.4585	95	-0.44	028.42	0
028	Manganese (mg / kg (ppm))	0038	158.00	2.0000	162.84	10.938	5.4585	95	-0.44	028.43	0
028	Manganese (mg / kg (ppm))	0039	158.00	1.6630	162.84	10.938	5.4585	95	-0.44	028.44	0
028	Manganese (mg / kg (ppm))	0358	158.32	13.100	162.84	10.938	5.4585	95	-0.41	028.41	0
028	Manganese (mg / kg (ppm))	0610	158.50	5.0000	162.84	10.938	5.4585	95	-0.40	028.43	0
028	Manganese (mg / kg (ppm))	2152	159.21	1.1600	162.84	10.938	5.4585	95	-0.33	028.44	0
028	Manganese (mg / kg (ppm))	0021	160.00	20.000	162.84	10.938	5.4585	95	-0.26	028.43	0
028	Manganese (mg / kg (ppm))	0033	160.00	10.000	162.84	10.938	5.4585	95	-0.26	028.43	0
028	Manganese (mg / kg (ppm))	0675	160.61	1.9000	162.84	10.938	5.4585	95	-0.20	028.43	0
028	Manganese (mg / kg (ppm))	0098	160.95	3.5000	162.84	10.938	5.4585	95	-0.17	028.41	0
028	Manganese (mg / kg (ppm))	0187	161.00	1.0000	162.84	10.938	5.4585	95	-0.17	028.42	0
028	Manganese (mg / kg (ppm))	0918	161.34	8.4100	162.84	10.938	5.4585	95	-0.14	028.43	0
028	Manganese (mg / kg (ppm))	0563	161.47	0.47820	162.84	10.938	5.4585	95	-0.13	028.31	0
028	Manganese (mg / kg (ppm))	0083	162.50	3.0000	162.84	10.938	5.4585	95	-0.03	028.43	0
028	Manganese (mg / kg (ppm))	2089	162.80	3.6100	162.84	10.938	5.4585	95	0.00	028.43	0
028	Manganese (mg / kg (ppm))	0366	163.00	18.000	162.84	10.938	5.4585	95	0.01	028.42	0
028	Manganese (mg / kg (ppm))	0572	163.00	6.0000	162.84	10.938	5.4585	95	0.01	028.53	0
028	Manganese (mg / kg (ppm))	0199	163.40	5.8000	162.84	10.938	5.4585	95	0.05	028.53	0
028	Manganese (mg / kg (ppm))	0425	163.95	4.9000	162.84	10.938	5.4585	95	0.10	028.43	0
028	Manganese (mg / kg (ppm))	0629	164.00	0.00000	162.84	10.938	5.4585	95	0.11	028.31	0
028	Manganese (mg / kg (ppm))	0529	164.13	4.7500	162.84	10.938	5.4585	95	0.12	028.31	0
028	Manganese (mg / kg (ppm))	0553	164.50	3.0000	162.84	10.938	5.4585	95	0.15	028.53	0
028	Manganese (mg / kg (ppm))	0074	165.00	14.000	162.84	10.938	5.4585	95	0.20	028.41	0
028	Manganese (mg / kg (ppm))	0939	165.20	1.8200	162.84	10.938	5.4585	95	0.22	028.31	0
028	Manganese (mg / kg (ppm))	0656	165.33	7.1300	162.84	10.938	5.4585	95	0.23	028.31	0
028	Manganese (mg / kg (ppm))	0038	165.50	7.0000	162.84	10.938	5.4585	95	0.24	028.32	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
028	Manganese (mg / kg (ppm))	0884	165.70	9.8000	162.84	10.938	5.4585	95	0.26	028.31	0
028	Manganese (mg / kg (ppm))	0910	166.00	0.00000	162.84	10.938	5.4585	95	0.29	028.41	0
028	Manganese (mg / kg (ppm))	0560	166.25	1.5000	162.84	10.938	5.4585	95	0.31	028.42	0
028	Manganese (mg / kg (ppm))	0208	166.45	10.700	162.84	10.938	5.4585	95	0.33	028.41	0
028	Manganese (mg / kg (ppm))	0045	166.50	1.0000	162.84	10.938	5.4585	95	0.33	028.42	0
028	Manganese (mg / kg (ppm))	0242	166.50	1.0000	162.84	10.938	5.4585	95	0.33	028.99	0
028	Manganese (mg / kg (ppm))	2129	166.50	3.0000	162.84	10.938	5.4585	95	0.33	028.99	0
028	Manganese (mg / kg (ppm))	0037	166.65	15.100	162.84	10.938	5.4585	95	0.35	028.42	0
028	Manganese (mg / kg (ppm))	0096	166.93	13.689	162.84	10.938	5.4585	95	0.37	028.52	0
028	Manganese (mg / kg (ppm))	0904	166.97	3.3700	162.84	10.938	5.4585	95	0.38	028.31	0
028	Manganese (mg / kg (ppm))	0035	167.00	0.00000	162.84	10.938	5.4585	95	0.38	028.43	0
028	Manganese (mg / kg (ppm))	0265	168.00	16.000	162.84	10.938	5.4585	95	0.47	028.42	0
028	Manganese (mg / kg (ppm))	0682	168.76	0.00000	162.84	10.938	5.4585	95	0.54	028.43	0
028	Manganese (mg / kg (ppm))	0229	169.14	9.2100	162.84	10.938	5.4585	95	0.58	028.41	0
028	Manganese (mg / kg (ppm))	0035	170.00	2.0000	162.84	10.938	5.4585	95	0.65	028.33	0
028	Manganese (mg / kg (ppm))	0878	170.50	1.0000	162.84	10.938	5.4585	95	0.70	028.31	0
028	Manganese (mg / kg (ppm))	0357	170.60	0.12900	162.84	10.938	5.4585	95	0.71	028.42	0
028	Manganese (mg / kg (ppm))	0536	171.14	0.08000	162.84	10.938	5.4585	95	0.76	028.31	0
028	Manganese (mg / kg (ppm))	0964	171.14	2.9330	162.84	10.938	5.4585	95	0.76	028.43	0
028	Manganese (mg / kg (ppm))	0629	172.00	0.00000	162.84	10.938	5.4585	95	0.84	028.43	0
028	Manganese (mg / kg (ppm))	0407	172.39	2.7988	162.84	10.938	5.4585	95	0.87	028.41	0
028	Manganese (mg / kg (ppm))	0190	173.00	3.0000	162.84	10.938	5.4585	95	0.93	028.42	0
028	Manganese (mg / kg (ppm))	0861	173.00	2.0000	162.84	10.938	5.4585	95	0.93	028.43	0
028	Manganese (mg / kg (ppm))	0354	173.02	3.0100	162.84	10.938	5.4585	95	0.93	028.31	0
028	Manganese (mg / kg (ppm))	0098	173.60	3.0000	162.84	10.938	5.4585	95	0.98	028.43	0
028	Manganese (mg / kg (ppm))	0726	173.65	0.30000	162.84	10.938	5.4585	95	0.99	028.42	0
028	Manganese (mg / kg (ppm))	0208	175.00	14.000	162.84	10.938	5.4585	95	1.11	028.31	0
028	Manganese (mg / kg (ppm))	0027	175.25	5.1040	162.84	10.938	5.4585	95	1.13	028.43	0
028	Manganese (mg / kg (ppm))	0692	176.00	20.000	162.84	10.938	5.4585	95	1.20	028.42	0
028	Manganese (mg / kg (ppm))	0100	176.00	8.0000	162.84	10.938	5.4585	95	1.20	028.99	0
028	Manganese (mg / kg (ppm))	0590	176.00	6.0000	162.84	10.938	5.4585	95	1.20	028.99	0
028	Manganese (mg / kg (ppm))	0066	178.16	4.7100	162.84	10.938	5.4585	95	1.40	028.31	0
028	Manganese (mg / kg (ppm))	0034	178.63	5.3400	162.84	10.938	5.4585	95	1.44	028.43	0
028	Manganese (mg / kg (ppm))	0175	179.00	2.0000	162.84	10.938	5.4585	95	1.48	028.31	0
028	Manganese (mg / kg (ppm))	0504	180.25	0.10000	162.84	10.938	5.4585	95	1.59	028.33	0
028	Manganese (mg / kg (ppm))	0504	185.10	2.0000	162.84	10.938	5.4585	95	2.03	028.32	0
028	Manganese (mg / kg (ppm))	0169	186.00	18.000	162.84	10.938	5.4585	95	2.12	028.43	0
028	Manganese (mg / kg (ppm))	0294	188.45	2.5000	162.84	10.938	5.4585	95	2.34	028.42	0
028	Manganese (mg / kg (ppm))	0186	231.70	28.000	162.84	10.938	5.4585	95	6.30	028.42	0
031	Phosphorus (%)	0405	0.05995	0.00090	0.60552	0.02927	0.55501	138	-18.64	031.41	0
031	Phosphorus (%)	0026	0.44500	0.03000	0.60552	0.02927	0.55501	138	-5.48	031.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	0536	0.47000	0.04000	0.60552	0.02927	0.55501	138	-4.63	031.06	0
031	Phosphorus (%)	0108	0.48500	0.01000	0.60552	0.02927	0.55501	138	-4.12	031.01	0
031	Phosphorus (%)	0017	0.49500	0.03000	0.60552	0.02927	0.55501	138	-3.78	031.43	0
031	Phosphorus (%)	2129	0.51100	0.00200	0.60552	0.02927	0.55501	138	-3.23	031.42	0
031	Phosphorus (%)	0018	0.52400	0.07600	0.60552	0.02927	0.55501	138	-2.79	031.01	0
031	Phosphorus (%)	0610	0.53000	0.00000	0.60552	0.02927	0.55501	138	-2.58	031.43	0
031	Phosphorus (%)	0940	0.55000	0.00000	0.60552	0.02927	0.55501	138	-1.90	031.01	0
031	Phosphorus (%)	0278	0.55000	0.00000	0.60552	0.02927	0.55501	138	-1.90	031.42	0
031	Phosphorus (%)	0874	0.55405	0.00290	0.60552	0.02927	0.55501	138	-1.76	031.01	0
031	Phosphorus (%)	0689	0.55500	0.01000	0.60552	0.02927	0.55501	138	-1.73	031.01	0
031	Phosphorus (%)	0187	0.55830	0.00060	0.60552	0.02927	0.55501	138	-1.61	031.42	0
031	Phosphorus (%)	0693	0.56150	0.02500	0.60552	0.02927	0.55501	138	-1.50	031.42	0
031	Phosphorus (%)	0100	0.56500	0.03000	0.60552	0.02927	0.55501	138	-1.38	031.99	0
031	Phosphorus (%)	0144	0.57000	0.00000	0.60552	0.02927	0.55501	138	-1.21	031.41	0
031	Phosphorus (%)	0848	0.57000	0.02000	0.60552	0.02927	0.55501	138	-1.21	031.41	0
031	Phosphorus (%)	0154	0.57125	0.04810	0.60552	0.02927	0.55501	138	-1.17	031.52	0
031	Phosphorus (%)	0021	0.57500	0.07000	0.60552	0.02927	0.55501	138	-1.04	031.43	0
031	Phosphorus (%)	0675	0.57500	0.01000	0.60552	0.02927	0.55501	138	-1.04	031.43	0
031	Phosphorus (%)	0620	0.57560	0.00280	0.60552	0.02927	0.55501	138	-1.02	031.01	0
031	Phosphorus (%)	0038	0.57650	0.00300	0.60552	0.02927	0.55501	138	-0.99	031.01	0
031	Phosphorus (%)	2009	0.57755	0.00270	0.60552	0.02927	0.55501	138	-0.96	031.01	0
031	Phosphorus (%)	0885	0.57860	0.00880	0.60552	0.02927	0.55501	138	-0.92	031.01	0
031	Phosphorus (%)	0009	0.57935	0.01710	0.60552	0.02927	0.55501	138	-0.89	031.42	0
031	Phosphorus (%)	0066	0.58000	0.00000	0.60552	0.02927	0.55501	138	-0.87	031.01	0
031	Phosphorus (%)	0035	0.58000	0.02000	0.60552	0.02927	0.55501	138	-0.87	031.43	0
031	Phosphorus (%)	2152	0.58000	0.00000	0.60552	0.02927	0.55501	138	-0.87	031.44	0
031	Phosphorus (%)	0910	0.58000	0.04000	0.60552	0.02927	0.55501	138	-0.87	031.41	0
031	Phosphorus (%)	0520	0.58200	0.00400	0.60552	0.02927	0.55501	138	-0.80	031.41	0
031	Phosphorus (%)	0038	0.58200	0.00400	0.60552	0.02927	0.55501	138	-0.80	031.43	0
031	Phosphorus (%)	0849	0.58250	0.00500	0.60552	0.02927	0.55501	138	-0.79	031.01	0
031	Phosphorus (%)	0529	0.58500	0.03000	0.60552	0.02927	0.55501	138	-0.70	031.01	0
031	Phosphorus (%)	0169	0.58500	0.01000	0.60552	0.02927	0.55501	138	-0.70	031.03	0
031	Phosphorus (%)	0294	0.58500	0.01000	0.60552	0.02927	0.55501	138	-0.70	031.42	0
031	Phosphorus (%)	0043	0.58500	0.01000	0.60552	0.02927	0.55501	138	-0.70	031.43	0
031	Phosphorus (%)	0233	0.59000	0.00000	0.60552	0.02927	0.55501	138	-0.53	031.01	0
031	Phosphorus (%)	0171	0.59000	0.00000	0.60552	0.02927	0.55501	138	-0.53	031.41	0
031	Phosphorus (%)	0298	0.59000	0.02000	0.60552	0.02927	0.55501	138	-0.53	031.41	0
031	Phosphorus (%)	0353	0.59000	0.02000	0.60552	0.02927	0.55501	138	-0.53	031.43	0
031	Phosphorus (%)	0629	0.59000	0.00000	0.60552	0.02927	0.55501	138	-0.53	031.43	0
031	Phosphorus (%)	0676	0.59000	0.00000	0.60552	0.02927	0.55501	138	-0.53	031.99	0
031	Phosphorus (%)	0039	0.59100	0.02200	0.60552	0.02927	0.55501	138	-0.50	031.44	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	2006	0.59300	0.00000	0.60552	0.02927	0.55501	138	-0.43	031.01	0
031	Phosphorus (%)	0175	0.59500	0.01000	0.60552	0.02927	0.55501	138	-0.36	031.01	0
031	Phosphorus (%)	0016	0.59750	0.01500	0.60552	0.02927	0.55501	138	-0.27	031.01	0
031	Phosphorus (%)	0553	0.59750	0.06300	0.60552	0.02927	0.55501	138	-0.27	031.53	0
031	Phosphorus (%)	0098	0.59835	0.01050	0.60552	0.02927	0.55501	138	-0.24	031.41	0
031	Phosphorus (%)	0001	0.59850	0.01100	0.60552	0.02927	0.55501	138	-0.24	031.03	0
031	Phosphorus (%)	0152	0.60000	0.02000	0.60552	0.02927	0.55501	138	-0.19	031.01	0
031	Phosphorus (%)	0629	0.60000	0.00000	0.60552	0.02927	0.55501	138	-0.19	031.01	0
031	Phosphorus (%)	0674	0.60000	0.00000	0.60552	0.02927	0.55501	138	-0.19	031.01	0
031	Phosphorus (%)	2022	0.60000	0.00000	0.60552	0.02927	0.55501	138	-0.19	031.01	0
031	Phosphorus (%)	2109	0.60000	0.03600	0.60552	0.02927	0.55501	138	-0.19	031.01	0
031	Phosphorus (%)	0013	0.60000	0.00000	0.60552	0.02927	0.55501	138	-0.19	031.02	0
031	Phosphorus (%)	0413	0.60000	0.00000	0.60552	0.02927	0.55501	138	-0.19	031.41	0
031	Phosphorus (%)	0265	0.60000	0.02000	0.60552	0.02927	0.55501	138	-0.19	031.42	0
031	Phosphorus (%)	0682	0.60000	0.00000	0.60552	0.02927	0.55501	138	-0.19	031.43	0
031	Phosphorus (%)	2089	0.60000	0.02000	0.60552	0.02927	0.55501	138	-0.19	031.43	0
031	Phosphorus (%)	0208	0.60050	0.03100	0.60552	0.02927	0.55501	138	-0.17	031.01	0
031	Phosphorus (%)	0142	0.60205	0.00350	0.60552	0.02927	0.55501	138	-0.12	031.01	0
031	Phosphorus (%)	0096	0.60330	0.01340	0.60552	0.02927	0.55501	138	-0.08	031.42	0
031	Phosphorus (%)	0504	0.60400	0.04200	0.60552	0.02927	0.55501	138	-0.05	031.42	0
031	Phosphorus (%)	0034	0.60400	0.01040	0.60552	0.02927	0.55501	138	-0.05	031.43	0
031	Phosphorus (%)	0148	0.60450	0.00300	0.60552	0.02927	0.55501	138	-0.03	031.41	0
031	Phosphorus (%)	0354	0.60500	0.01000	0.60552	0.02927	0.55501	138	-0.02	031.01	0
031	Phosphorus (%)	0505	0.60500	0.01000	0.60552	0.02927	0.55501	138	-0.02	031.02	0
031	Phosphorus (%)	0504	0.60500	0.01000	0.60552	0.02927	0.55501	138	-0.02	031.03	0
031	Phosphorus (%)	0229	0.60500	0.01000	0.60552	0.02927	0.55501	138	-0.02	031.41	0
031	Phosphorus (%)	0968	0.60500	0.00000	0.60552	0.02927	0.55501	138	-0.02	031.43	0
031	Phosphorus (%)	0186	0.60735	0.02450	0.60552	0.02927	0.55501	138	0.06	031.42	0
031	Phosphorus (%)	0878	0.60750	0.00100	0.60552	0.02927	0.55501	138	0.07	031.01	0
031	Phosphorus (%)	0626	0.60850	0.00700	0.60552	0.02927	0.55501	138	0.10	031.01	0
031	Phosphorus (%)	0357	0.60915	0.02150	0.60552	0.02927	0.55501	138	0.12	031.42	0
031	Phosphorus (%)	0590	0.61000	0.02000	0.60552	0.02927	0.55501	138	0.15	031.01	0
031	Phosphorus (%)	0019	0.61000	0.00000	0.60552	0.02927	0.55501	138	0.15	031.41	0
031	Phosphorus (%)	0049	0.61000	0.00000	0.60552	0.02927	0.55501	138	0.15	031.41	0
031	Phosphorus (%)	0089	0.61000	0.00000	0.60552	0.02927	0.55501	138	0.15	031.41	0
031	Phosphorus (%)	2012	0.61000	0.02000	0.60552	0.02927	0.55501	138	0.15	031.41	0
031	Phosphorus (%)	0242	0.61000	0.00000	0.60552	0.02927	0.55501	138	0.15	031.99	0
031	Phosphorus (%)	0560	0.61160	0.01000	0.60552	0.02927	0.55501	138	0.21	031.42	0
031	Phosphorus (%)	0139	0.61300	0.00000	0.60552	0.02927	0.55501	138	0.26	031.41	0
031	Phosphorus (%)	0861	0.61350	0.01300	0.60552	0.02927	0.55501	138	0.27	031.43	0
031	Phosphorus (%)	0036	0.61350	0.03020	0.60552	0.02927	0.55501	138	0.27	031.03	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	2062	0.61485	0.00770	0.60552	0.02927	0.55501	138	0.32	031.01	0
031	Phosphorus (%)	0650	0.61500	0.03200	0.60552	0.02927	0.55501	138	0.32	031.01	0
031	Phosphorus (%)	0897	0.61500	0.01000	0.60552	0.02927	0.55501	138	0.32	031.01	0
031	Phosphorus (%)	0939	0.61500	0.03000	0.60552	0.02927	0.55501	138	0.32	031.01	0
031	Phosphorus (%)	0043	0.61500	0.01000	0.60552	0.02927	0.55501	138	0.32	031.02	0
031	Phosphorus (%)	0164	0.61500	0.01000	0.60552	0.02927	0.55501	138	0.32	031.41	0
031	Phosphorus (%)	0358	0.61500	0.03000	0.60552	0.02927	0.55501	138	0.32	031.41	0
031	Phosphorus (%)	0083	0.61500	0.01000	0.60552	0.02927	0.55501	138	0.32	031.43	0
031	Phosphorus (%)	0425	0.61500	0.01000	0.60552	0.02927	0.55501	138	0.32	031.43	0
031	Phosphorus (%)	0205	0.61600	0.00000	0.60552	0.02927	0.55501	138	0.36	031.42	0
031	Phosphorus (%)	0138	0.61700	0.00800	0.60552	0.02927	0.55501	138	0.39	031.06	0
031	Phosphorus (%)	0563	0.61870	0.00120	0.60552	0.02927	0.55501	138	0.45	031.01	0
031	Phosphorus (%)	0619	0.61950	0.00900	0.60552	0.02927	0.55501	138	0.48	031.41	0
031	Phosphorus (%)	0646	0.62000	0.00000	0.60552	0.02927	0.55501	138	0.49	031.01	0
031	Phosphorus (%)	0948	0.62000	0.02000	0.60552	0.02927	0.55501	138	0.49	031.01	0
031	Phosphorus (%)	0263	0.62045	0.00090	0.60552	0.02927	0.55501	138	0.51	031.42	0
031	Phosphorus (%)	0345	0.62050	0.00700	0.60552	0.02927	0.55501	138	0.51	031.43	0
031	Phosphorus (%)	0893	0.62500	0.01000	0.60552	0.02927	0.55501	138	0.67	031.00	0
031	Phosphorus (%)	0895	0.62500	0.01000	0.60552	0.02927	0.55501	138	0.67	031.00	0
031	Phosphorus (%)	0511	0.62500	0.09000	0.60552	0.02927	0.55501	138	0.67	031.01	0
031	Phosphorus (%)	0884	0.62500	0.01000	0.60552	0.02927	0.55501	138	0.67	031.01	0
031	Phosphorus (%)	0903	0.62500	0.01000	0.60552	0.02927	0.55501	138	0.67	031.01	0
031	Phosphorus (%)	2053	0.62500	0.01000	0.60552	0.02927	0.55501	138	0.67	031.02	0
031	Phosphorus (%)	0037	0.62500	0.01000	0.60552	0.02927	0.55501	138	0.67	031.42	0
031	Phosphorus (%)	0510	0.62500	0.01000	0.60552	0.02927	0.55501	138	0.67	031.43	0
031	Phosphorus (%)	0199	0.62500	0.01000	0.60552	0.02927	0.55501	138	0.67	031.53	0
031	Phosphorus (%)	0014	0.62550	0.03100	0.60552	0.02927	0.55501	138	0.68	031.42	0
031	Phosphorus (%)	0918	0.62700	0.01200	0.60552	0.02927	0.55501	138	0.73	031.43	0
031	Phosphorus (%)	0904	0.63000	0.02000	0.60552	0.02927	0.55501	138	0.84	031.01	0
031	Phosphorus (%)	0300	0.63150	0.06900	0.60552	0.02927	0.55501	138	0.89	031.41	0
031	Phosphorus (%)	0027	0.63200	0.02000	0.60552	0.02927	0.55501	138	0.90	031.43	0
031	Phosphorus (%)	0889	0.63200	0.00800	0.60552	0.02927	0.55501	138	0.90	031.99	0
031	Phosphorus (%)	0552	0.63500	0.01000	0.60552	0.02927	0.55501	138	1.01	031.01	0
031	Phosphorus (%)	0074	0.63500	0.01000	0.60552	0.02927	0.55501	138	1.01	031.41	0
031	Phosphorus (%)	0028	0.63500	0.01000	0.60552	0.02927	0.55501	138	1.01	031.43	0
031	Phosphorus (%)	0572	0.63550	0.05100	0.60552	0.02927	0.55501	138	1.02	031.53	0
031	Phosphorus (%)	0726	0.63640	0.00020	0.60552	0.02927	0.55501	138	1.06	031.42	0
031	Phosphorus (%)	0036	0.63825	0.00730	0.60552	0.02927	0.55501	138	1.12	031.43	0
031	Phosphorus (%)	0190	0.64000	0.02000	0.60552	0.02927	0.55501	138	1.18	031.42	0
031	Phosphorus (%)	0033	0.64500	0.01000	0.60552	0.02927	0.55501	138	1.35	031.43	0
031	Phosphorus (%)	0598	0.64755	0.00750	0.60552	0.02927	0.55501	138	1.44	031.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	0692	0.65000	0.04000	0.60552	0.02927	0.55501	138	1.52	031.42	0
031	Phosphorus (%)	0045	0.65100	0.01600	0.60552	0.02927	0.55501	138	1.55	031.42	0
031	Phosphorus (%)	0512	0.65160	0.03260	0.60552	0.02927	0.55501	138	1.57	031.43	0
031	Phosphorus (%)	0852	0.65500	0.01000	0.60552	0.02927	0.55501	138	1.69	031.99	0
031	Phosphorus (%)	0964	0.65545	0.01890	0.60552	0.02927	0.55501	138	1.71	031.43	0
031	Phosphorus (%)	0366	0.66000	0.02000	0.60552	0.02927	0.55501	138	1.86	031.42	0
031	Phosphorus (%)	0226	0.66500	0.01000	0.60552	0.02927	0.55501	138	2.03	031.41	0
031	Phosphorus (%)	0613	0.66500	0.07000	0.60552	0.02927	0.55501	138	2.03	031.42	0
031	Phosphorus (%)	0098	0.66700	0.00640	0.60552	0.02927	0.55501	138	2.10	031.43	0
031	Phosphorus (%)	0407	0.66880	0.02740	0.60552	0.02927	0.55501	138	2.16	031.41	0
031	Phosphorus (%)	0208	0.67300	0.02800	0.60552	0.02927	0.55501	138	2.31	031.41	0
031	Phosphorus (%)	2128	0.79550	0.00900	0.60552	0.02927	0.55501	138	6.49	031.01	0
031	Phosphorus (%)	0202	5,274.3	74.520	0.60552	0.02927	0.55501	138	180172.71	031.43	0
032	Potassium (%)	0405	0.06330	0.00080	0.67804	0.03821	0.70406	94	-16.09	032.41	0
032	Potassium (%)	2022	0.59000	0.00800	0.67804	0.03821	0.70406	94	-2.30	032.31	0
032	Potassium (%)	0017	0.59000	0.04000	0.67804	0.03821	0.70406	94	-2.30	032.43	0
032	Potassium (%)	0536	0.60000	0.00000	0.67804	0.03821	0.70406	94	-2.04	032.31	0
032	Potassium (%)	0035	0.61500	0.01000	0.67804	0.03821	0.70406	94	-1.65	032.43	0
032	Potassium (%)	0144	0.62000	0.02000	0.67804	0.03821	0.70406	94	-1.52	032.41	0
032	Potassium (%)	0049	0.63000	0.04000	0.67804	0.03821	0.70406	94	-1.26	032.41	0
032	Potassium (%)	0169	0.63250	0.00300	0.67804	0.03821	0.70406	94	-1.19	032.32	0
032	Potassium (%)	0511	0.63500	0.03000	0.67804	0.03821	0.70406	94	-1.13	032.41	0
032	Potassium (%)	0510	0.63500	0.01000	0.67804	0.03821	0.70406	94	-1.13	032.43	0
032	Potassium (%)	0874	0.63900	0.00200	0.67804	0.03821	0.70406	94	-1.02	032.31	0
032	Potassium (%)	0354	0.64000	0.00000	0.67804	0.03821	0.70406	94	-1.00	032.31	0
032	Potassium (%)	0294	0.64000	0.02000	0.67804	0.03821	0.70406	94	-1.00	032.42	0
032	Potassium (%)	0629	0.64000	0.00000	0.67804	0.03821	0.70406	94	-1.00	032.43	0
032	Potassium (%)	0142	0.64055	0.00070	0.67804	0.03821	0.70406	94	-0.98	032.31	0
032	Potassium (%)	0175	0.64500	0.01000	0.67804	0.03821	0.70406	94	-0.86	032.31	0
032	Potassium (%)	0904	0.64500	0.01000	0.67804	0.03821	0.70406	94	-0.86	032.31	0
032	Potassium (%)	0019	0.64500	0.01000	0.67804	0.03821	0.70406	94	-0.86	032.41	0
032	Potassium (%)	0610	0.64500	0.01000	0.67804	0.03821	0.70406	94	-0.86	032.43	0
032	Potassium (%)	0038	0.64600	0.01000	0.67804	0.03821	0.70406	94	-0.84	032.43	0
032	Potassium (%)	0154	0.64735	0.05790	0.67804	0.03821	0.70406	94	-0.80	032.52	0
032	Potassium (%)	0021	0.65000	0.02000	0.67804	0.03821	0.70406	94	-0.73	032.43	0
032	Potassium (%)	0878	0.65100	0.00200	0.67804	0.03821	0.70406	94	-0.71	032.31	0
032	Potassium (%)	0505	0.65250	0.00500	0.67804	0.03821	0.70406	94	-0.67	032.31	0
032	Potassium (%)	0263	0.65300	0.00300	0.67804	0.03821	0.70406	94	-0.66	032.32	0
032	Potassium (%)	0529	0.65500	0.09000	0.67804	0.03821	0.70406	94	-0.60	032.31	0
032	Potassium (%)	0650	0.65500	0.03000	0.67804	0.03821	0.70406	94	-0.60	032.31	0
032	Potassium (%)	0358	0.65500	0.01000	0.67804	0.03821	0.70406	94	-0.60	032.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
032	Potassium (%)	0038	0.65900	0.00800	0.67804	0.03821	0.70406	94	-0.50	032.31	0
032	Potassium (%)	0413	0.66000	0.02000	0.67804	0.03821	0.70406	94	-0.47	032.41	0
032	Potassium (%)	2012	0.66000	0.00000	0.67804	0.03821	0.70406	94	-0.47	032.41	0
032	Potassium (%)	0613	0.66000	0.04000	0.67804	0.03821	0.70406	94	-0.47	032.42	0
032	Potassium (%)	0096	0.66110	0.02200	0.67804	0.03821	0.70406	94	-0.44	032.42	0
032	Potassium (%)	0598	0.66150	0.01340	0.67804	0.03821	0.70406	94	-0.43	032.41	0
032	Potassium (%)	0504	0.66500	0.01000	0.67804	0.03821	0.70406	94	-0.34	032.02	0
032	Potassium (%)	0629	0.66500	0.01000	0.67804	0.03821	0.70406	94	-0.34	032.31	0
032	Potassium (%)	0039	0.66650	0.01900	0.67804	0.03821	0.70406	94	-0.30	032.44	0
032	Potassium (%)	0148	0.66800	0.00800	0.67804	0.03821	0.70406	94	-0.26	032.41	0
032	Potassium (%)	0861	0.66800	0.01600	0.67804	0.03821	0.70406	94	-0.26	032.43	0
032	Potassium (%)	0229	0.67000	0.02000	0.67804	0.03821	0.70406	94	-0.21	032.41	0
032	Potassium (%)	0278	0.67000	0.02000	0.67804	0.03821	0.70406	94	-0.21	032.42	0
032	Potassium (%)	0619	0.67250	0.00100	0.67804	0.03821	0.70406	94	-0.15	032.41	0
032	Potassium (%)	0208	0.67300	0.02000	0.67804	0.03821	0.70406	94	-0.13	032.41	0
032	Potassium (%)	0918	0.67450	0.00700	0.67804	0.03821	0.70406	94	-0.09	032.43	0
032	Potassium (%)	0164	0.67500	0.01000	0.67804	0.03821	0.70406	94	-0.08	032.41	0
032	Potassium (%)	0910	0.67500	0.03000	0.67804	0.03821	0.70406	94	-0.08	032.41	0
032	Potassium (%)	0100	0.67500	0.01000	0.67804	0.03821	0.70406	94	-0.08	032.99	0
032	Potassium (%)	0345	0.67550	0.01100	0.67804	0.03821	0.70406	94	-0.07	032.43	0
032	Potassium (%)	0553	0.67550	0.02900	0.67804	0.03821	0.70406	94	-0.07	032.53	0
032	Potassium (%)	0171	0.67600	0.01400	0.67804	0.03821	0.70406	94	-0.05	032.41	0
032	Potassium (%)	0689	0.68000	0.02000	0.67804	0.03821	0.70406	94	0.05	032.31	0
032	Potassium (%)	0682	0.68000	0.00000	0.67804	0.03821	0.70406	94	0.05	032.43	0
032	Potassium (%)	0357	0.68005	0.01750	0.67804	0.03821	0.70406	94	0.05	032.42	0
032	Potassium (%)	0693	0.68100	0.00600	0.67804	0.03821	0.70406	94	0.08	032.42	0
032	Potassium (%)	0034	0.68195	0.00570	0.67804	0.03821	0.70406	94	0.10	032.43	0
032	Potassium (%)	2053	0.68500	0.01000	0.67804	0.03821	0.70406	94	0.18	032.31	0
032	Potassium (%)	0037	0.68500	0.03000	0.67804	0.03821	0.70406	94	0.18	032.42	0
032	Potassium (%)	0083	0.68500	0.01000	0.67804	0.03821	0.70406	94	0.18	032.43	0
032	Potassium (%)	0675	0.68500	0.01000	0.67804	0.03821	0.70406	94	0.18	032.43	0
032	Potassium (%)	0242	0.68500	0.01000	0.67804	0.03821	0.70406	94	0.18	032.99	0
032	Potassium (%)	0205	0.68800	0.00600	0.67804	0.03821	0.70406	94	0.26	032.42	0
032	Potassium (%)	0035	0.69000	0.02000	0.67804	0.03821	0.70406	94	0.31	032.33	0
032	Potassium (%)	0074	0.69000	0.00000	0.67804	0.03821	0.70406	94	0.31	032.41	0
032	Potassium (%)	0098	0.69000	0.02000	0.67804	0.03821	0.70406	94	0.31	032.43	0
032	Potassium (%)	0199	0.69000	0.00000	0.67804	0.03821	0.70406	94	0.31	032.53	0
032	Potassium (%)	0045	0.69200	0.03000	0.67804	0.03821	0.70406	94	0.37	032.42	0
032	Potassium (%)	0187	0.69350	0.00040	0.67804	0.03821	0.70406	94	0.40	032.42	0
032	Potassium (%)	0425	0.69500	0.01000	0.67804	0.03821	0.70406	94	0.44	032.43	0
032	Potassium (%)	0139	0.69500	0.02400	0.67804	0.03821	0.70406	94	0.44	032.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
032	Potassium (%)	0560	0.69935	0.01630	0.67804	0.03821	0.70406	94	0.56	032.42	0
032	Potassium (%)	0939	0.70000	0.00000	0.67804	0.03821	0.70406	94	0.57	032.31	0
032	Potassium (%)	0043	0.70200	0.00200	0.67804	0.03821	0.70406	94	0.63	032.43	0
032	Potassium (%)	0504	0.70400	0.01400	0.67804	0.03821	0.70406	94	0.68	032.42	0
032	Potassium (%)	0098	0.70500	0.01000	0.67804	0.03821	0.70406	94	0.71	032.41	0
032	Potassium (%)	0208	0.70850	0.00100	0.67804	0.03821	0.70406	94	0.80	032.31	0
032	Potassium (%)	0948	0.71500	0.01000	0.67804	0.03821	0.70406	94	0.97	032.31	0
032	Potassium (%)	0353	0.71500	0.01000	0.67804	0.03821	0.70406	94	0.97	032.43	0
032	Potassium (%)	0009	0.71755	0.00710	0.67804	0.03821	0.70406	94	1.03	032.42	0
032	Potassium (%)	0186	0.72085	0.02970	0.67804	0.03821	0.70406	94	1.12	032.42	0
032	Potassium (%)	0572	0.72250	0.00900	0.67804	0.03821	0.70406	94	1.16	032.53	0
032	Potassium (%)	0563	0.72375	0.01970	0.67804	0.03821	0.70406	94	1.20	032.31	0
032	Potassium (%)	0520	0.72650	0.00900	0.67804	0.03821	0.70406	94	1.27	032.41	0
032	Potassium (%)	0726	0.73005	0.00250	0.67804	0.03821	0.70406	94	1.36	032.42	0
032	Potassium (%)	0968	0.73400	0.00400	0.67804	0.03821	0.70406	94	1.46	032.43	0
032	Potassium (%)	0265	0.73500	0.09000	0.67804	0.03821	0.70406	94	1.49	032.42	0
032	Potassium (%)	0964	0.74175	0.03970	0.67804	0.03821	0.70406	94	1.67	032.43	0
032	Potassium (%)	0066	0.74500	0.01000	0.67804	0.03821	0.70406	94	1.75	032.31	0
032	Potassium (%)	0692	0.75500	0.07000	0.67804	0.03821	0.70406	94	2.01	032.42	0
032	Potassium (%)	0407	0.75640	0.01740	0.67804	0.03821	0.70406	94	2.05	032.41	0
032	Potassium (%)	0190	0.77500	0.01000	0.67804	0.03821	0.70406	94	2.54	032.42	0
032	Potassium (%)	2089	0.78000	0.04000	0.67804	0.03821	0.70406	94	2.67	032.43	0
032	Potassium (%)	0226	0.83000	0.00000	0.67804	0.03821	0.70406	94	3.98	032.41	0
032	Potassium (%)	0366	0.89000	0.02000	0.67804	0.03821	0.70406	94	5.55	032.42	0
032	Potassium (%)	0202	6,233.8	64.720	0.67804	0.03821	0.70406	94	163120.11	032.43	0
032	Potassium (%)	0889	0.00000	0.00000	0.67804	0.03821	0.70406	94		032.99	4
033	Salt as chloride (%)	0539	0.13600	0.00000	0.32206	0.04887	0.01944	67	-3.81	033.00	0
033	Salt as chloride (%)	0726	0.15600	0.00000	0.32206	0.04887	0.01944	67	-3.40	033.03	0
033	Salt as chloride (%)	0904	0.18500	0.01000	0.32206	0.04887	0.01944	67	-2.80	033.05	0
033	Salt as chloride (%)	0265	0.22000	0.02000	0.32206	0.04887	0.01944	67	-2.09	033.03	0
033	Salt as chloride (%)	0511	0.23000	0.06000	0.32206	0.04887	0.01944	67	-1.88	033.00	0
033	Salt as chloride (%)	0893	0.26000	0.00000	0.32206	0.04887	0.01944	67	-1.27	033.00	0
033	Salt as chloride (%)	0144	0.26000	0.02000	0.32206	0.04887	0.01944	67	-1.27	033.03	0
033	Salt as chloride (%)	0366	0.26500	0.01000	0.32206	0.04887	0.01944	67	-1.17	033.00	0
033	Salt as chloride (%)	0171	0.26500	0.01000	0.32206	0.04887	0.01944	67	-1.17	033.05	0
033	Salt as chloride (%)	0298	0.27000	0.02000	0.32206	0.04887	0.01944	67	-1.07	033.00	0
033	Salt as chloride (%)	0674	0.27000	0.00000	0.32206	0.04887	0.01944	67	-1.07	033.00	0
033	Salt as chloride (%)	0045	0.27400	0.01200	0.32206	0.04887	0.01944	67	-0.98	033.00	0
033	Salt as chloride (%)	0190	0.28000	0.04000	0.32206	0.04887	0.01944	67	-0.86	033.03	0
033	Salt as chloride (%)	0504	0.28200	0.01800	0.32206	0.04887	0.01944	67	-0.82	033.00	0
033	Salt as chloride (%)	0505	0.29000	0.02000	0.32206	0.04887	0.01944	67	-0.66	033.03	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
033	Salt as chloride (%)	0208	0.29400	0.02400	0.32206	0.04887	0.01944	67	-0.57	033.99	0
033	Salt as chloride (%)	0083	0.29500	0.01000	0.32206	0.04887	0.01944	67	-0.55	033.01	0
033	Salt as chloride (%)	2073	0.29500	0.03000	0.32206	0.04887	0.01944	67	-0.55	033.99	0
033	Salt as chloride (%)	0619	0.29800	0.01000	0.32206	0.04887	0.01944	67	-0.49	033.01	0
033	Salt as chloride (%)	2062	0.29815	0.00410	0.32206	0.04887	0.01944	67	-0.49	033.00	0
033	Salt as chloride (%)	0878	0.30000	0.00000	0.32206	0.04887	0.01944	67	-0.45	033.05	0
033	Salt as chloride (%)	0098	0.30500	0.03000	0.32206	0.04887	0.01944	67	-0.35	033.01	0
033	Salt as chloride (%)	0948	0.30500	0.01000	0.32206	0.04887	0.01944	67	-0.35	033.01	0
033	Salt as chloride (%)	0689	0.30500	0.01000	0.32206	0.04887	0.01944	67	-0.35	033.05	0
033	Salt as chloride (%)	0693	0.30900	0.03000	0.32206	0.04887	0.01944	67	-0.27	033.00	0
033	Salt as chloride (%)	0169	0.31000	0.00000	0.32206	0.04887	0.01944	67	-0.25	033.00	0
033	Salt as chloride (%)	0884	0.31000	0.00000	0.32206	0.04887	0.01944	67	-0.25	033.00	0
033	Salt as chloride (%)	0650	0.31000	0.00000	0.32206	0.04887	0.01944	67	-0.25	033.01	0
033	Salt as chloride (%)	0910	0.31000	0.00000	0.32206	0.04887	0.01944	67	-0.25	033.99	0
033	Salt as chloride (%)	0895	0.31500	0.01000	0.32206	0.04887	0.01944	67	-0.14	033.00	0
033	Salt as chloride (%)	0903	0.31500	0.01000	0.32206	0.04887	0.01944	67	-0.14	033.00	0
033	Salt as chloride (%)	0148	0.31850	0.00100	0.32206	0.04887	0.01944	67	-0.07	033.01	0
033	Salt as chloride (%)	0407	0.31850	0.00100	0.32206	0.04887	0.01944	67	-0.07	033.01	0
033	Salt as chloride (%)	0354	0.32000	0.00000	0.32206	0.04887	0.01944	67	-0.04	033.01	0
033	Salt as chloride (%)	0001	0.32500	0.01000	0.32206	0.04887	0.01944	67	0.06	033.01	0
033	Salt as chloride (%)	2109	0.32500	0.01000	0.32206	0.04887	0.01944	67	0.06	033.01	0
033	Salt as chloride (%)	0205	0.32700	0.01200	0.32206	0.04887	0.01944	67	0.10	033.01	0
033	Salt as chloride (%)	0885	0.32815	0.00850	0.32206	0.04887	0.01944	67	0.12	033.00	0
033	Salt as chloride (%)	0242	0.33000	0.00000	0.32206	0.04887	0.01944	67	0.16	033.01	0
033	Salt as chloride (%)	0425	0.33000	0.00000	0.32206	0.04887	0.01944	67	0.16	033.01	0
033	Salt as chloride (%)	0510	0.33000	0.00000	0.32206	0.04887	0.01944	67	0.16	033.01	0
033	Salt as chloride (%)	0629	0.33000	0.00000	0.32206	0.04887	0.01944	67	0.16	033.01	0
033	Salt as chloride (%)	0897	0.33500	0.01000	0.32206	0.04887	0.01944	67	0.26	033.00	0
033	Salt as chloride (%)	0199	0.33500	0.01000	0.32206	0.04887	0.01944	67	0.26	033.01	0
033	Salt as chloride (%)	0226	0.34000	0.02000	0.32206	0.04887	0.01944	67	0.37	033.01	0
033	Salt as chloride (%)	0590	0.34000	0.04000	0.32206	0.04887	0.01944	67	0.37	033.01	0
033	Salt as chloride (%)	0164	0.34000	0.00000	0.32206	0.04887	0.01944	67	0.37	033.01	0
033	Salt as chloride (%)	2009	0.34410	0.00180	0.32206	0.04887	0.01944	67	0.45	033.00	0
033	Salt as chloride (%)	0964	0.34450	0.00100	0.32206	0.04887	0.01944	67	0.46	033.01	0
033	Salt as chloride (%)	0861	0.34500	0.01000	0.32206	0.04887	0.01944	67	0.47	033.01	0
033	Salt as chloride (%)	0175	0.35000	0.02000	0.32206	0.04887	0.01944	67	0.57	033.01	0
033	Salt as chloride (%)	0644	0.35000	0.00000	0.32206	0.04887	0.01944	67	0.57	033.01	0
033	Salt as chloride (%)	0682	0.35000	0.00000	0.32206	0.04887	0.01944	67	0.57	033.99	0
033	Salt as chloride (%)	0536	0.35500	0.01000	0.32206	0.04887	0.01944	67	0.67	033.99	0
033	Salt as chloride (%)	0874	0.36290	0.00000	0.32206	0.04887	0.01944	67	0.84	033.01	0
033	Salt as chloride (%)	0100	0.36500	0.03000	0.32206	0.04887	0.01944	67	0.88	033.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
033	Salt as chloride (%)	0552	0.37000	0.04000	0.32206	0.04887	0.01944	67	0.98	033.99	0
033	Salt as chloride (%)	0353	0.38000	0.02000	0.32206	0.04887	0.01944	67	1.19	033.00	0
033	Salt as chloride (%)	0613	0.39270	0.16940	0.32206	0.04887	0.01944	67	1.45	033.99	0
033	Salt as chloride (%)	0003	0.40000	0.00000	0.32206	0.04887	0.01944	67	1.59	033.01	0
033	Salt as chloride (%)	0017	0.44500	0.07000	0.32206	0.04887	0.01944	67	2.52	033.99	0
033	Salt as chloride (%)	0019	0.45000	0.00000	0.32206	0.04887	0.01944	67	2.62	033.01	0
033	Salt as chloride (%)	0358	0.52000	0.04000	0.32206	0.04887	0.01944	67	4.05	033.99	0
033	Salt as chloride (%)	0921	0.52500	0.07000	0.32206	0.04887	0.01944	67	4.15	033.00	0
033	Salt as chloride (%)	0278	0.53500	0.27000	0.32206	0.04887	0.01944	67	4.36	033.01	0
033	Salt as chloride (%)	0027	0.55700	0.00000	0.32206	0.04887	0.01944	67	4.81	033.99	0
033	Salt as chloride (%)	0675	0.57500	0.01000	0.32206	0.04887	0.01944	67	5.18	033.00	0
034	Selenium (mg / kg (ppm))	0619	0.40700	0.01600	0.67383	0.17827	0.03634	23	-1.50	034.41	0
034	Selenium (mg / kg (ppm))	0964	0.42670	0.00700	0.67383	0.17827	0.03634	23	-1.39	034.43	0
034	Selenium (mg / kg (ppm))	0199	0.53000	0.00000	0.67383	0.17827	0.03634	23	-0.81	034.53	0
034	Selenium (mg / kg (ppm))	0164	0.55500	0.01000	0.67383	0.17827	0.03634	23	-0.67	034.04	0
034	Selenium (mg / kg (ppm))	0171	0.58000	0.00000	0.67383	0.17827	0.03634	23	-0.53	034.04	0
034	Selenium (mg / kg (ppm))	2053	0.58500	0.01000	0.67383	0.17827	0.03634	23	-0.50	034.04	0
034	Selenium (mg / kg (ppm))	0169	0.59350	0.00300	0.67383	0.17827	0.03634	23	-0.45	034.04	0
034	Selenium (mg / kg (ppm))	0096	0.59465	0.01930	0.67383	0.17827	0.03634	23	-0.44	034.52	0
034	Selenium (mg / kg (ppm))	0563	0.60075	0.11850	0.67383	0.17827	0.03634	23	-0.41	034.04	0
034	Selenium (mg / kg (ppm))	0021	0.61500	0.05000	0.67383	0.17827	0.03634	23	-0.33	034.53	0
034	Selenium (mg / kg (ppm))	0038	0.62700	0.02000	0.67383	0.17827	0.03634	23	-0.26	034.01	0
034	Selenium (mg / kg (ppm))	0910	0.63000	0.02000	0.67383	0.17827	0.03634	23	-0.25	034.52	0
034	Selenium (mg / kg (ppm))	0098	0.63500	0.01000	0.67383	0.17827	0.03634	23	-0.22	034.01	0
034	Selenium (mg / kg (ppm))	0553	0.63800	0.02400	0.67383	0.17827	0.03634	23	-0.20	034.53	0
034	Selenium (mg / kg (ppm))	0610	0.66000	0.04000	0.67383	0.17827	0.03634	23	-0.08	034.04	0
034	Selenium (mg / kg (ppm))	0208	0.69800	0.04400	0.67383	0.17827	0.03634	23	0.14	034.52	0
034	Selenium (mg / kg (ppm))	0918	0.71760	0.01300	0.67383	0.17827	0.03634	23	0.25	034.53	0
034	Selenium (mg / kg (ppm))	0939	0.81500	0.03000	0.67383	0.17827	0.03634	23	0.79	034.04	0
034	Selenium (mg / kg (ppm))	0675	0.82500	0.01000	0.67383	0.17827	0.03634	23	0.85	034.43	0
034	Selenium (mg / kg (ppm))	0098	1.0195	0.05700	0.67383	0.17827	0.03634	23	1.94	034.53	0
034	Selenium (mg / kg (ppm))	0560	1.3360	0.00400	0.67383	0.17827	0.03634	23	3.71	034.52	0
034	Selenium (mg / kg (ppm))	0692	1.9000	0.00000	0.67383	0.17827	0.03634	23	6.88	034.42	0
034	Selenium (mg / kg (ppm))	0190	2.4650	0.33000	0.67383	0.17827	0.03634	23	10.05	034.42	0
034	Selenium (mg / kg (ppm))	0021	0.00000	0.00000	0.67383	0.17827	0.03634	23		034.43	4
035	Sodium (%)	0405	0.02025	0.00090	0.20875	0.01554	0.33633	105	-12.13	035.41	0
035	Sodium (%)	0878	0.16000	0.00000	0.20875	0.01554	0.33633	105	-3.14	035.01	0
035	Sodium (%)	0017	0.17500	0.03000	0.20875	0.01554	0.33633	105	-2.17	035.43	0
035	Sodium (%)	0043	0.17800	0.00000	0.20875	0.01554	0.33633	105	-1.98	035.43	0
035	Sodium (%)	0021	0.18000	0.00000	0.20875	0.01554	0.33633	105	-1.85	035.43	0
035	Sodium (%)	0175	0.18500	0.01000	0.20875	0.01554	0.33633	105	-1.53	035.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
035	Sodium (%)	0035	0.18500	0.01000	0.20875	0.01554	0.33633	105	-1.53	035.43	0
035	Sodium (%)	0242	0.18500	0.01000	0.20875	0.01554	0.33633	105	-1.53	035.99	0
035	Sodium (%)	0512	0.18695	0.00390	0.20875	0.01554	0.33633	105	-1.40	035.43	0
035	Sodium (%)	0142	0.18875	0.00390	0.20875	0.01554	0.33633	105	-1.29	035.31	0
035	Sodium (%)	0294	0.19000	0.00000	0.20875	0.01554	0.33633	105	-1.21	035.42	0
035	Sodium (%)	0610	0.19000	0.00000	0.20875	0.01554	0.33633	105	-1.21	035.43	0
035	Sodium (%)	0154	0.19100	0.00380	0.20875	0.01554	0.33633	105	-1.14	035.52	0
035	Sodium (%)	0861	0.19200	0.00400	0.20875	0.01554	0.33633	105	-1.08	035.43	0
035	Sodium (%)	0038	0.19300	0.01200	0.20875	0.01554	0.33633	105	-1.01	035.43	0
035	Sodium (%)	0014	0.19400	0.00000	0.20875	0.01554	0.33633	105	-0.95	035.31	0
035	Sodium (%)	0354	0.19500	0.01000	0.20875	0.01554	0.33633	105	-0.89	035.31	0
035	Sodium (%)	0144	0.19500	0.01000	0.20875	0.01554	0.33633	105	-0.89	035.41	0
035	Sodium (%)	0510	0.19550	0.00300	0.20875	0.01554	0.33633	105	-0.85	035.43	0
035	Sodium (%)	0169	0.19800	0.00600	0.20875	0.01554	0.33633	105	-0.69	035.32	0
035	Sodium (%)	0263	0.19820	0.00160	0.20875	0.01554	0.33633	105	-0.68	035.32	0
035	Sodium (%)	0096	0.19845	0.00870	0.20875	0.01554	0.33633	105	-0.66	035.42	0
035	Sodium (%)	0034	0.19935	0.02150	0.20875	0.01554	0.33633	105	-0.61	035.43	0
035	Sodium (%)	0884	0.20000	0.00000	0.20875	0.01554	0.33633	105	-0.56	035.02	0
035	Sodium (%)	0035	0.20000	0.02000	0.20875	0.01554	0.33633	105	-0.56	035.33	0
035	Sodium (%)	0019	0.20000	0.02000	0.20875	0.01554	0.33633	105	-0.56	035.41	0
035	Sodium (%)	0229	0.20000	0.00000	0.20875	0.01554	0.33633	105	-0.56	035.41	0
035	Sodium (%)	0298	0.20000	0.02000	0.20875	0.01554	0.33633	105	-0.56	035.41	0
035	Sodium (%)	0300	0.20000	0.00200	0.20875	0.01554	0.33633	105	-0.56	035.41	0
035	Sodium (%)	2012	0.20000	0.00000	0.20875	0.01554	0.33633	105	-0.56	035.41	0
035	Sodium (%)	0278	0.20000	0.00000	0.20875	0.01554	0.33633	105	-0.56	035.42	0
035	Sodium (%)	0098	0.20000	0.00000	0.20875	0.01554	0.33633	105	-0.56	035.43	0
035	Sodium (%)	0199	0.20000	0.00000	0.20875	0.01554	0.33633	105	-0.56	035.53	0
035	Sodium (%)	0205	0.20300	0.00000	0.20875	0.01554	0.33633	105	-0.37	035.42	0
035	Sodium (%)	2129	0.20350	0.00100	0.20875	0.01554	0.33633	105	-0.34	035.99	0
035	Sodium (%)	2006	0.20450	0.00700	0.20875	0.01554	0.33633	105	-0.27	035.01	0
035	Sodium (%)	0148	0.20450	0.00300	0.20875	0.01554	0.33633	105	-0.27	035.41	0
035	Sodium (%)	0693	0.20450	0.00300	0.20875	0.01554	0.33633	105	-0.27	035.42	0
035	Sodium (%)	0066	0.20500	0.01000	0.20875	0.01554	0.33633	105	-0.24	035.05	0
035	Sodium (%)	0152	0.20500	0.01000	0.20875	0.01554	0.33633	105	-0.24	035.05	0
035	Sodium (%)	0650	0.20500	0.01000	0.20875	0.01554	0.33633	105	-0.24	035.31	0
035	Sodium (%)	0904	0.20500	0.01000	0.20875	0.01554	0.33633	105	-0.24	035.31	0
035	Sodium (%)	0003	0.20500	0.01000	0.20875	0.01554	0.33633	105	-0.24	035.41	0
035	Sodium (%)	0358	0.20500	0.01000	0.20875	0.01554	0.33633	105	-0.24	035.41	0
035	Sodium (%)	0413	0.20500	0.01000	0.20875	0.01554	0.33633	105	-0.24	035.41	0
035	Sodium (%)	0520	0.20500	0.00400	0.20875	0.01554	0.33633	105	-0.24	035.41	0
035	Sodium (%)	0910	0.20500	0.01000	0.20875	0.01554	0.33633	105	-0.24	035.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
035	Sodium (%)	0265	0.20500	0.01000	0.20875	0.01554	0.33633	105	-0.24	035.42	0
035	Sodium (%)	0613	0.20500	0.01000	0.20875	0.01554	0.33633	105	-0.24	035.42	0
035	Sodium (%)	0425	0.20500	0.01000	0.20875	0.01554	0.33633	105	-0.24	035.43	0
035	Sodium (%)	0918	0.20650	0.00100	0.20875	0.01554	0.33633	105	-0.15	035.43	0
035	Sodium (%)	0039	0.20665	0.00670	0.20875	0.01554	0.33633	105	-0.14	035.41	0
035	Sodium (%)	0345	0.20700	0.00200	0.20875	0.01554	0.33633	105	-0.11	035.43	0
035	Sodium (%)	0208	0.20700	0.01200	0.20875	0.01554	0.33633	105	-0.11	035.31	0
035	Sodium (%)	0598	0.20745	0.00190	0.20875	0.01554	0.33633	105	-0.08	035.41	0
035	Sodium (%)	0164	0.20800	0.00400	0.20875	0.01554	0.33633	105	-0.05	035.41	0
035	Sodium (%)	0045	0.20800	0.00600	0.20875	0.01554	0.33633	105	-0.05	035.42	0
035	Sodium (%)	0563	0.20875	0.00550	0.20875	0.01554	0.33633	105	0.00	035.31	0
035	Sodium (%)	0572	0.20900	0.01000	0.20875	0.01554	0.33633	105	0.02	035.53	0
035	Sodium (%)	0504	0.20950	0.00900	0.20875	0.01554	0.33633	105	0.05	035.42	0
035	Sodium (%)	0233	0.21000	0.00000	0.20875	0.01554	0.33633	105	0.08	035.31	0
035	Sodium (%)	0939	0.21000	0.00000	0.20875	0.01554	0.33633	105	0.08	035.31	0
035	Sodium (%)	0049	0.21000	0.00000	0.20875	0.01554	0.33633	105	0.08	035.41	0
035	Sodium (%)	0089	0.21000	0.00000	0.20875	0.01554	0.33633	105	0.08	035.41	0
035	Sodium (%)	0682	0.21000	0.00000	0.20875	0.01554	0.33633	105	0.08	035.43	0
035	Sodium (%)	0353	0.21000	0.02000	0.20875	0.01554	0.33633	105	0.08	035.43	0
035	Sodium (%)	0560	0.21035	0.00090	0.20875	0.01554	0.33633	105	0.10	035.42	0
035	Sodium (%)	0878	0.21050	0.00100	0.20875	0.01554	0.33633	105	0.11	035.31	0
035	Sodium (%)	0096	0.21150	0.00140	0.20875	0.01554	0.33633	105	0.18	035.52	0
035	Sodium (%)	0138	0.21450	0.00500	0.20875	0.01554	0.33633	105	0.37	035.01	0
035	Sodium (%)	0098	0.21500	0.01000	0.20875	0.01554	0.33633	105	0.40	035.41	0
035	Sodium (%)	0171	0.21500	0.01000	0.20875	0.01554	0.33633	105	0.40	035.41	0
035	Sodium (%)	0190	0.21500	0.01000	0.20875	0.01554	0.33633	105	0.40	035.42	0
035	Sodium (%)	0083	0.21500	0.01000	0.20875	0.01554	0.33633	105	0.40	035.43	0
035	Sodium (%)	0100	0.21500	0.01000	0.20875	0.01554	0.33633	105	0.40	035.99	0
035	Sodium (%)	0948	0.21500	0.03000	0.20875	0.01554	0.33633	105	0.40	035.31	0
035	Sodium (%)	2109	0.21600	0.00400	0.20875	0.01554	0.33633	105	0.47	035.01	0
035	Sodium (%)	0553	0.21650	0.02700	0.20875	0.01554	0.33633	105	0.50	035.53	0
035	Sodium (%)	0874	0.21750	0.00700	0.20875	0.01554	0.33633	105	0.56	035.31	0
035	Sodium (%)	0504	0.22000	0.00000	0.20875	0.01554	0.33633	105	0.72	035.05	0
035	Sodium (%)	0629	0.22000	0.00000	0.20875	0.01554	0.33633	105	0.72	035.31	0
035	Sodium (%)	0186	0.22215	0.01570	0.20875	0.01554	0.33633	105	0.86	035.42	0
035	Sodium (%)	0726	0.22265	0.00350	0.20875	0.01554	0.33633	105	0.89	035.42	0
035	Sodium (%)	0038	0.22300	0.01000	0.20875	0.01554	0.33633	105	0.92	035.31	0
035	Sodium (%)	0208	0.22300	0.01800	0.20875	0.01554	0.33633	105	0.92	035.41	0
035	Sodium (%)	0536	0.22500	0.01000	0.20875	0.01554	0.33633	105	1.05	035.31	0
035	Sodium (%)	2053	0.22500	0.01000	0.20875	0.01554	0.33633	105	1.05	035.42	0
035	Sodium (%)	2089	0.22500	0.01000	0.20875	0.01554	0.33633	105	1.05	035.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
035	Sodium (%)	0590	0.22500	0.01000	0.20875	0.01554	0.33633	105	1.05	035.99	0
035	Sodium (%)	0505	0.22700	0.00600	0.20875	0.01554	0.33633	105	1.17	035.31	0
035	Sodium (%)	0187	0.22705	0.00330	0.20875	0.01554	0.33633	105	1.18	035.42	0
035	Sodium (%)	0629	0.22800	0.00400	0.20875	0.01554	0.33633	105	1.24	035.43	0
035	Sodium (%)	0689	0.23000	0.02000	0.20875	0.01554	0.33633	105	1.37	035.31	0
035	Sodium (%)	0407	0.23025	0.00230	0.20875	0.01554	0.33633	105	1.38	035.41	0
035	Sodium (%)	0692	0.23500	0.01000	0.20875	0.01554	0.33633	105	1.69	035.42	0
035	Sodium (%)	0366	0.24000	0.02000	0.20875	0.01554	0.33633	105	2.01	035.42	0
035	Sodium (%)	0033	0.24000	0.00000	0.20875	0.01554	0.33633	105	2.01	035.43	0
035	Sodium (%)	0619	0.24450	0.00900	0.20875	0.01554	0.33633	105	2.30	035.41	0
035	Sodium (%)	0675	0.24500	0.01000	0.20875	0.01554	0.33633	105	2.33	035.43	0
035	Sodium (%)	0226	0.24600	0.00800	0.20875	0.01554	0.33633	105	2.40	035.41	0
035	Sodium (%)	0139	0.25150	0.00300	0.20875	0.01554	0.33633	105	2.75	035.41	0
035	Sodium (%)	0964	0.25290	0.00680	0.20875	0.01554	0.33633	105	2.84	035.43	0
035	Sodium (%)	0529	0.25450	0.00100	0.20875	0.01554	0.33633	105	2.94	035.31	0
035	Sodium (%)	0889	0.42000	0.08000	0.20875	0.01554	0.33633	105	13.60	035.99	0
035	Sodium (%)	0202	1,940.0	34.500	0.20875	0.01554	0.33633	105	124851.39	035.43	0
036	Sulfur (%)	0098	0.33000	0.02000	0.37762	0.03257	7.5974	41	-1.46	036.04	0
036	Sulfur (%)	0229	0.33000	0.00000	0.37762	0.03257	7.5974	41	-1.46	036.04	0
036	Sulfur (%)	0610	0.33000	0.00000	0.37762	0.03257	7.5974	41	-1.46	036.04	0
036	Sulfur (%)	0366	0.33500	0.01000	0.37762	0.03257	7.5974	41	-1.31	036.42	0
036	Sulfur (%)	0096	0.34255	0.00310	0.37762	0.03257	7.5974	41	-1.08	036.42	0
036	Sulfur (%)	0038	0.34400	0.03200	0.37762	0.03257	7.5974	41	-1.03	036.43	0
036	Sulfur (%)	0693	0.34950	0.00100	0.37762	0.03257	7.5974	41	-0.86	036.42	0
036	Sulfur (%)	0357	0.35465	0.01110	0.37762	0.03257	7.5974	41	-0.71	036.42	0
036	Sulfur (%)	2089	0.35500	0.01000	0.37762	0.03257	7.5974	41	-0.69	036.43	0
036	Sulfur (%)	0407	0.35655	0.00050	0.37762	0.03257	7.5974	41	-0.65	036.42	0
036	Sulfur (%)	0345	0.35700	0.00600	0.37762	0.03257	7.5974	41	-0.63	036.43	0
036	Sulfur (%)	0164	0.36000	0.00000	0.37762	0.03257	7.5974	41	-0.54	036.42	0
036	Sulfur (%)	0035	0.36000	0.02000	0.37762	0.03257	7.5974	41	-0.54	036.43	0
036	Sulfur (%)	0045	0.36350	0.02500	0.37762	0.03257	7.5974	41	-0.43	036.42	0
036	Sulfur (%)	0629	0.36350	0.00100	0.37762	0.03257	7.5974	41	-0.43	036.43	0
036	Sulfur (%)	0278	0.36500	0.01000	0.37762	0.03257	7.5974	41	-0.39	036.42	0
036	Sulfur (%)	0187	0.36750	0.00080	0.37762	0.03257	7.5974	41	-0.31	036.42	0
036	Sulfur (%)	0171	0.36900	0.00200	0.37762	0.03257	7.5974	41	-0.26	036.42	0
036	Sulfur (%)	0353	0.37000	0.04000	0.37762	0.03257	7.5974	41	-0.23	036.43	0
036	Sulfur (%)	0038	0.37050	0.00500	0.37762	0.03257	7.5974	41	-0.22	036.00	0
036	Sulfur (%)	0708	0.37300	0.00400	0.37762	0.03257	7.5974	41	-0.14	036.42	0
036	Sulfur (%)	0226	0.37500	0.03000	0.37762	0.03257	7.5974	41	-0.08	036.04	0
036	Sulfur (%)	0021	0.38000	0.04000	0.37762	0.03257	7.5974	41	0.07	036.43	0
036	Sulfur (%)	0918	0.38150	0.00900	0.37762	0.03257	7.5974	41	0.12	036.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
036	Sulfur (%)	0726	0.38350	0.00160	0.37762	0.03257	7.5974	41	0.18	036.42	0
036	Sulfur (%)	0098	0.38490	0.00680	0.37762	0.03257	7.5974	41	0.22	036.43	0
036	Sulfur (%)	0613	0.38500	0.01000	0.37762	0.03257	7.5974	41	0.23	036.42	0
036	Sulfur (%)	0242	0.39000	0.00000	0.37762	0.03257	7.5974	41	0.38	036.99	0
036	Sulfur (%)	0560	0.39485	0.00410	0.37762	0.03257	7.5974	41	0.53	036.42	0
036	Sulfur (%)	0190	0.39500	0.01000	0.37762	0.03257	7.5974	41	0.53	036.42	0
036	Sulfur (%)	0169	0.39500	0.01000	0.37762	0.03257	7.5974	41	0.53	036.43	0
036	Sulfur (%)	0889	0.39800	0.00200	0.37762	0.03257	7.5974	41	0.63	036.99	0
036	Sulfur (%)	0964	0.40235	0.00050	0.37762	0.03257	7.5974	41	0.76	036.43	0
036	Sulfur (%)	0692	0.41000	0.02000	0.37762	0.03257	7.5974	41	0.99	036.42	0
036	Sulfur (%)	0037	0.41500	0.01000	0.37762	0.03257	7.5974	41	1.15	036.42	0
036	Sulfur (%)	0265	0.42000	0.00000	0.37762	0.03257	7.5974	41	1.30	036.42	0
036	Sulfur (%)	0510	0.42000	0.02000	0.37762	0.03257	7.5974	41	1.30	036.43	0
036	Sulfur (%)	0294	0.44000	0.00000	0.37762	0.03257	7.5974	41	1.92	036.42	0
036	Sulfur (%)	0186	0.44915	0.01970	0.37762	0.03257	7.5974	41	2.20	036.42	0
036	Sulfur (%)	0202	3,269.7	37.100	0.37762	0.03257	7.5974	41	100372.78	036.43	0
036	Sulfur (%)	0910	3,547.0	274.00	0.37762	0.03257	7.5974	41	108887.61	036.52	0
037	Zinc (mg / kg (ppm))	0405	10.600	0.30000	135.21	10.560	5.7952	100	-11.80	037.41	0
037	Zinc (mg / kg (ppm))	0536	28.060	1.4200	135.21	10.560	5.7952	100	-10.15	037.31	0
037	Zinc (mg / kg (ppm))	0026	86.500	9.0000	135.21	10.560	5.7952	100	-4.61	037.42	0
037	Zinc (mg / kg (ppm))	0009	104.47	3.0900	135.21	10.560	5.7952	100	-2.91	037.42	0
037	Zinc (mg / kg (ppm))	0017	112.90	2.9100	135.21	10.560	5.7952	100	-2.11	037.43	0
037	Zinc (mg / kg (ppm))	0278	117.15	0.90000	135.21	10.560	5.7952	100	-1.71	037.42	0
037	Zinc (mg / kg (ppm))	0889	117.50	3.0000	135.21	10.560	5.7952	100	-1.68	037.99	0
037	Zinc (mg / kg (ppm))	0202	118.25	4.1600	135.21	10.560	5.7952	100	-1.61	037.43	0
037	Zinc (mg / kg (ppm))	0675	120.61	4.4500	135.21	10.560	5.7952	100	-1.38	037.43	0
037	Zinc (mg / kg (ppm))	0629	121.00	0.00000	135.21	10.560	5.7952	100	-1.35	037.43	0
037	Zinc (mg / kg (ppm))	0164	122.50	1.0000	135.21	10.560	5.7952	100	-1.20	037.41	0
037	Zinc (mg / kg (ppm))	0504	122.85	10.500	135.21	10.560	5.7952	100	-1.17	037.33	0
037	Zinc (mg / kg (ppm))	0353	123.00	4.0000	135.21	10.560	5.7952	100	-1.16	037.43	0
037	Zinc (mg / kg (ppm))	0357	123.72	0.32300	135.21	10.560	5.7952	100	-1.09	037.42	0
037	Zinc (mg / kg (ppm))	0358	124.84	6.2200	135.21	10.560	5.7952	100	-0.98	037.41	0
037	Zinc (mg / kg (ppm))	0021	125.00	30.000	135.21	10.560	5.7952	100	-0.97	037.43	0
037	Zinc (mg / kg (ppm))	0345	126.50	3.0000	135.21	10.560	5.7952	100	-0.82	037.43	0
037	Zinc (mg / kg (ppm))	0689	126.60	2.6000	135.21	10.560	5.7952	100	-0.81	037.31	0
037	Zinc (mg / kg (ppm))	0096	126.94	21.215	135.21	10.560	5.7952	100	-0.78	037.42	0
037	Zinc (mg / kg (ppm))	0171	127.50	1.0000	135.21	10.560	5.7952	100	-0.73	037.41	0
037	Zinc (mg / kg (ppm))	0918	127.84	0.21000	135.21	10.560	5.7952	100	-0.70	037.43	0
037	Zinc (mg / kg (ppm))	2053	128.00	2.0000	135.21	10.560	5.7952	100	-0.68	037.42	0
037	Zinc (mg / kg (ppm))	0553	128.00	10.000	135.21	10.560	5.7952	100	-0.68	037.53	0
037	Zinc (mg / kg (ppm))	0910	128.50	11.000	135.21	10.560	5.7952	100	-0.63	037.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
037	Zinc (mg / kg (ppm))	0199	128.65	7.1000	135.21	10.560	5.7952	100	-0.62	037.53	0
037	Zinc (mg / kg (ppm))	0187	129.08	0.55000	135.21	10.560	5.7952	100	-0.58	037.42	0
037	Zinc (mg / kg (ppm))	0520	129.45	5.3000	135.21	10.560	5.7952	100	-0.55	037.41	0
037	Zinc (mg / kg (ppm))	0242	129.50	3.0000	135.21	10.560	5.7952	100	-0.54	037.99	0
037	Zinc (mg / kg (ppm))	0590	129.50	1.0000	135.21	10.560	5.7952	100	-0.54	037.99	0
037	Zinc (mg / kg (ppm))	0039	129.80	2.3980	135.21	10.560	5.7952	100	-0.51	037.44	0
037	Zinc (mg / kg (ppm))	2022	130.00	2.0000	135.21	10.560	5.7952	100	-0.49	037.31	0
037	Zinc (mg / kg (ppm))	0610	130.00	6.0000	135.21	10.560	5.7952	100	-0.49	037.43	0
037	Zinc (mg / kg (ppm))	0049	130.10	3.2900	135.21	10.560	5.7952	100	-0.48	037.41	0
037	Zinc (mg / kg (ppm))	0038	130.50	5.0000	135.21	10.560	5.7952	100	-0.45	037.43	0
037	Zinc (mg / kg (ppm))	0083	131.00	6.0000	135.21	10.560	5.7952	100	-0.40	037.43	0
037	Zinc (mg / kg (ppm))	2152	131.37	0.18000	135.21	10.560	5.7952	100	-0.36	037.44	0
037	Zinc (mg / kg (ppm))	0904	131.52	1.0400	135.21	10.560	5.7952	100	-0.35	037.31	0
037	Zinc (mg / kg (ppm))	0066	131.63	3.1100	135.21	10.560	5.7952	100	-0.34	037.31	0
037	Zinc (mg / kg (ppm))	0939	131.93	1.3000	135.21	10.560	5.7952	100	-0.31	037.31	0
037	Zinc (mg / kg (ppm))	0510	132.00	8.0000	135.21	10.560	5.7952	100	-0.30	037.43	0
037	Zinc (mg / kg (ppm))	2089	132.23	2.4500	135.21	10.560	5.7952	100	-0.28	037.43	0
037	Zinc (mg / kg (ppm))	0033	132.50	9.0000	135.21	10.560	5.7952	100	-0.26	037.43	0
037	Zinc (mg / kg (ppm))	0425	132.70	22.600	135.21	10.560	5.7952	100	-0.24	037.43	0
037	Zinc (mg / kg (ppm))	0001	133.50	7.0000	135.21	10.560	5.7952	100	-0.16	037.31	0
037	Zinc (mg / kg (ppm))	0613	133.65	8.5000	135.21	10.560	5.7952	100	-0.15	037.42	0
037	Zinc (mg / kg (ppm))	0560	133.75	0.50000	135.21	10.560	5.7952	100	-0.14	037.42	0
037	Zinc (mg / kg (ppm))	0964	133.81	7.4720	135.21	10.560	5.7952	100	-0.13	037.43	0
037	Zinc (mg / kg (ppm))	0878	134.00	0.00000	135.21	10.560	5.7952	100	-0.11	037.31	0
037	Zinc (mg / kg (ppm))	0098	134.10	7.0000	135.21	10.560	5.7952	100	-0.10	037.41	0
037	Zinc (mg / kg (ppm))	0010	135.00	10.000	135.21	10.560	5.7952	100	-0.02	037.33	0
037	Zinc (mg / kg (ppm))	0598	135.10	9.8000	135.21	10.560	5.7952	100	-0.01	037.41	0
037	Zinc (mg / kg (ppm))	0407	135.37	8.7685	135.21	10.560	5.7952	100	0.02	037.41	0
037	Zinc (mg / kg (ppm))	0148	135.38	0.55000	135.21	10.560	5.7952	100	0.02	037.41	0
037	Zinc (mg / kg (ppm))	0019	135.50	23.000	135.21	10.560	5.7952	100	0.03	037.41	0
037	Zinc (mg / kg (ppm))	0693	135.50	3.0000	135.21	10.560	5.7952	100	0.03	037.42	0
037	Zinc (mg / kg (ppm))	2012	135.84	9.8200	135.21	10.560	5.7952	100	0.06	037.41	0
037	Zinc (mg / kg (ppm))	0096	136.31	4.4202	135.21	10.560	5.7952	100	0.10	037.52	0
037	Zinc (mg / kg (ppm))	2149	136.38	1.7800	135.21	10.560	5.7952	100	0.11	037.31	0
037	Zinc (mg / kg (ppm))	0074	136.50	9.0000	135.21	10.560	5.7952	100	0.12	037.41	0
037	Zinc (mg / kg (ppm))	0563	137.00	4.0000	135.21	10.560	5.7952	100	0.17	037.31	0
037	Zinc (mg / kg (ppm))	0265	137.00	8.0000	135.21	10.560	5.7952	100	0.17	037.42	0
037	Zinc (mg / kg (ppm))	0100	137.00	22.000	135.21	10.560	5.7952	100	0.17	037.99	0
037	Zinc (mg / kg (ppm))	0045	137.50	11.000	135.21	10.560	5.7952	100	0.22	037.42	0
037	Zinc (mg / kg (ppm))	0028	137.50	3.0000	135.21	10.560	5.7952	100	0.22	037.43	0
037	Zinc (mg / kg (ppm))	0043	138.00	6.0000	135.21	10.560	5.7952	100	0.26	037.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
037	Zinc (mg / kg (ppm))	0968	138.50	1.0000	135.21	10.560	5.7952	100	0.31	037.43	0
037	Zinc (mg / kg (ppm))	0572	138.50	7.0000	135.21	10.560	5.7952	100	0.31	037.53	0
037	Zinc (mg / kg (ppm))	0366	139.50	11.000	135.21	10.560	5.7952	100	0.41	037.42	0
037	Zinc (mg / kg (ppm))	2129	140.50	3.0000	135.21	10.560	5.7952	100	0.50	037.99	0
037	Zinc (mg / kg (ppm))	0208	140.60	6.2000	135.21	10.560	5.7952	100	0.51	037.41	0
037	Zinc (mg / kg (ppm))	0656	140.81	0.76000	135.21	10.560	5.7952	100	0.53	037.31	0
037	Zinc (mg / kg (ppm))	0505	141.50	13.000	135.21	10.560	5.7952	100	0.60	037.31	0
037	Zinc (mg / kg (ppm))	0682	141.94	0.00000	135.21	10.560	5.7952	100	0.64	037.43	0
037	Zinc (mg / kg (ppm))	0511	142.00	6.0000	135.21	10.560	5.7952	100	0.64	037.41	0
037	Zinc (mg / kg (ppm))	0861	142.00	4.0000	135.21	10.560	5.7952	100	0.64	037.43	0
037	Zinc (mg / kg (ppm))	0504	142.25	5.3000	135.21	10.560	5.7952	100	0.67	037.32	0
037	Zinc (mg / kg (ppm))	0884	142.35	8.9000	135.21	10.560	5.7952	100	0.68	037.31	0
037	Zinc (mg / kg (ppm))	0208	142.50	7.0000	135.21	10.560	5.7952	100	0.69	037.31	0
037	Zinc (mg / kg (ppm))	0354	142.53	3.8000	135.21	10.560	5.7952	100	0.69	037.31	0
037	Zinc (mg / kg (ppm))	0874	143.00	10.000	135.21	10.560	5.7952	100	0.74	037.31	0
037	Zinc (mg / kg (ppm))	0038	144.00	2.0000	135.21	10.560	5.7952	100	0.83	037.32	0
037	Zinc (mg / kg (ppm))	0098	144.45	0.50000	135.21	10.560	5.7952	100	0.88	037.43	0
037	Zinc (mg / kg (ppm))	0948	144.53	0.27400	135.21	10.560	5.7952	100	0.88	037.33	0
037	Zinc (mg / kg (ppm))	0619	145.00	4.0000	135.21	10.560	5.7952	100	0.93	037.41	0
037	Zinc (mg / kg (ppm))	0190	146.35	0.70000	135.21	10.560	5.7952	100	1.06	037.42	0
037	Zinc (mg / kg (ppm))	2062	147.35	11.500	135.21	10.560	5.7952	100	1.15	037.31	0
037	Zinc (mg / kg (ppm))	0169	147.50	3.0000	135.21	10.560	5.7952	100	1.16	037.43	0
037	Zinc (mg / kg (ppm))	0037	149.55	1.3000	135.21	10.560	5.7952	100	1.36	037.42	0
037	Zinc (mg / kg (ppm))	0726	149.55	0.90000	135.21	10.560	5.7952	100	1.36	037.42	0
037	Zinc (mg / kg (ppm))	0529	150.00	2.5000	135.21	10.560	5.7952	100	1.40	037.31	0
037	Zinc (mg / kg (ppm))	0226	150.00	14.000	135.21	10.560	5.7952	100	1.40	037.41	0
037	Zinc (mg / kg (ppm))	0413	150.00	8.0000	135.21	10.560	5.7952	100	1.40	037.42	0
037	Zinc (mg / kg (ppm))	0294	152.83	1.2600	135.21	10.560	5.7952	100	1.67	037.42	0
037	Zinc (mg / kg (ppm))	0229	154.53	9.4200	135.21	10.560	5.7952	100	1.83	037.41	0
037	Zinc (mg / kg (ppm))	0692	155.00	24.000	135.21	10.560	5.7952	100	1.87	037.42	0
037	Zinc (mg / kg (ppm))	0027	158.14	13.478	135.21	10.560	5.7952	100	2.17	037.43	0
037	Zinc (mg / kg (ppm))	0512	161.45	1.5000	135.21	10.560	5.7952	100	2.49	037.43	0
037	Zinc (mg / kg (ppm))	0035	164.50	1.0000	135.21	10.560	5.7952	100	2.77	037.43	0
037	Zinc (mg / kg (ppm))	0035	182.00	2.0000	135.21	10.560	5.7952	100	4.43	037.33	0
037	Zinc (mg / kg (ppm))	0175	192.50	5.0000	135.21	10.560	5.7952	100	5.43	037.31	0
037	Zinc (mg / kg (ppm))	0186	0.00000	0.00000	135.21	10.560	5.7952	100		037.42	4
038	Molybdenum (mg / kg (ppm))	0353	0.67165	0.05550	1.1456	0.32642	0.11483	18	-1.45	038.43	0
038	Molybdenum (mg / kg (ppm))	0098	0.76350	0.02300	1.1456	0.32642	0.11483	18	-1.17	038.53	0
038	Molybdenum (mg / kg (ppm))	0910	0.79000	0.04000	1.1456	0.32642	0.11483	18	-1.09	038.52	0
038	Molybdenum (mg / kg (ppm))	0345	0.80000	0.02000	1.1456	0.32642	0.11483	18	-1.06	038.43	0
038	Molybdenum (mg / kg (ppm))	0918	0.91395	0.00390	1.1456	0.32642	0.11483	18	-0.71	038.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
038	Molybdenum (mg / kg (ppm))	0171	0.95000	0.10000	1.1456	0.32642	0.11483	18	-0.60	038.41	0
038	Molybdenum (mg / kg (ppm))	0169	1.0150	0.27000	1.1456	0.32642	0.11483	18	-0.40	038.43	0
038	Molybdenum (mg / kg (ppm))	2053	1.0700	0.02000	1.1456	0.32642	0.11483	18	-0.23	038.42	0
038	Molybdenum (mg / kg (ppm))	0964	1.1453	0.10070	1.1456	0.32642	0.11483	18	0.00	038.43	0
038	Molybdenum (mg / kg (ppm))	0226	1.2300	0.10000	1.1456	0.32642	0.11483	18	0.26	038.41	0
038	Molybdenum (mg / kg (ppm))	0096	1.2618	0.07390	1.1456	0.32642	0.11483	18	0.36	038.52	0
038	Molybdenum (mg / kg (ppm))	0278	1.2900	0.16000	1.1456	0.32642	0.11483	18	0.44	038.42	0
038	Molybdenum (mg / kg (ppm))	0038	1.3000	0.20000	1.1456	0.32642	0.11483	18	0.47	038.43	0
038	Molybdenum (mg / kg (ppm))	0553	1.3300	0.26000	1.1456	0.32642	0.11483	18	0.56	038.53	0
038	Molybdenum (mg / kg (ppm))	0693	1.4650	0.43000	1.1456	0.32642	0.11483	18	0.98	038.42	0
038	Molybdenum (mg / kg (ppm))	0021	1.5000	0.00000	1.1456	0.32642	0.11483	18	1.09	038.53	0
038	Molybdenum (mg / kg (ppm))	0045	1.5250	0.01000	1.1456	0.32642	0.11483	18	1.16	038.42	0
038	Molybdenum (mg / kg (ppm))	0510	1.6000	0.20000	1.1456	0.32642	0.11483	18	1.39	038.43	0
038	Molybdenum (mg / kg (ppm))	0021	0.00000	0.00000	1.1456	0.32642	0.11483	18		038.43	4
040	Barium (mg / kg (ppm))	0918	6.0996	0.17190				3		040.53	0
040	Barium (mg / kg (ppm))	0560	6.5950	0.19000				3		040.42	0
040	Barium (mg / kg (ppm))	0096	6.6097	0.15800				3		040.52	0
041	Vanadium (mg / kg (ppm))	0553	0.90800	0.05800				2		041.53	0
041	Vanadium (mg / kg (ppm))	0021	1.0850	0.23000				2		041.53	0
041	Vanadium (mg / kg (ppm))	0021	0.00000	0.00000				2		041.43	4
042	Chloride (%)	0413	0.14500	0.01000				3		042.02	0
042	Chloride (%)	0948	0.18500	0.01000				3		042.00	0
042	Chloride (%)	0613	0.25500	0.11000				3		042.99	0
101	Choline Chloride (mg / kg (ppm))	0644	570.00	4.0000				2		101.02	0
101	Choline Chloride (mg / kg (ppm))	0227	1,635.0	110.00				2		101.02	0
102	Niacin (mg / kg (ppm))	0675	89.575	2.1500				2		102.02	0
102	Niacin (mg / kg (ppm))	0227	181.50	1.0000				2		102.01	0
103	Pantothenic Acid (mg / kg (ppm))	0629	10.280	0.00000				2		103.02	0
103	Pantothenic Acid (mg / kg (ppm))	0227	24.600	0.40000				2		103.01	0
104	Riboflavin (mg / kg (ppm))	0910	21.205	1.3100	27.345	4.6048	1.4847	7	-1.33	104.03	0
104	Riboflavin (mg / kg (ppm))	0027	24.878	2.0931	27.345	4.6048	1.4847	7	-0.54	104.03	0
104	Riboflavin (mg / kg (ppm))	0208	25.100	0.80000	27.345	4.6048	1.4847	7	-0.49	104.03	0
104	Riboflavin (mg / kg (ppm))	0904	26.185	1.2900	27.345	4.6048	1.4847	7	-0.25	104.03	0
104	Riboflavin (mg / kg (ppm))	0227	30.950	2.7000	27.345	4.6048	1.4847	7	0.78	104.00	0
104	Riboflavin (mg / kg (ppm))	2103	31.100	0.60000	27.345	4.6048	1.4847	7	0.82	104.03	0
104	Riboflavin (mg / kg (ppm))	0171	32.000	1.6000	27.345	4.6048	1.4847	7	1.01	104.00	0
105	Thiamine (mg / kg (ppm))	0904	4.8250	0.73000	6.8208	2.2125	0.68250	6	-0.90	105.00	0
105	Thiamine (mg / kg (ppm))	0208	4.9795	1.7050	6.8208	2.2125	0.68250	6	-0.83	105.00	0
105	Thiamine (mg / kg (ppm))	0910	5.8350	0.91000	6.8208	2.2125	0.68250	6	-0.45	105.00	0
105	Thiamine (mg / kg (ppm))	0227	7.0300	0.60000	6.8208	2.2125	0.68250	6	0.09	105.01	0
105	Thiamine (mg / kg (ppm))	0629	9.0900	0.00000	6.8208	2.2125	0.68250	6	1.03	105.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
105	Thiamine (mg / kg (ppm))	2053	9.1650	0.15000	6.8208	2.2125	0.68250	6	1.06	105.00	0
106	Vitamin A (KU / kg)	0563	13.469	1.0580	30.835	6.0519	2.9569	29	-2.87	106.02	0
106	Vitamin A (KU / kg)	2068	22.500	1.0000	30.835	6.0519	2.9569	29	-1.38	106.02	0
106	Vitamin A (KU / kg)	0905	23.177	3.3450	30.835	6.0519	2.9569	29	-1.27	106.02	0
106	Vitamin A (KU / kg)	2053	25.300	3.2000	30.835	6.0519	2.9569	29	-0.91	106.02	0
106	Vitamin A (KU / kg)	0904	26.135	0.39000	30.835	6.0519	2.9569	29	-0.78	106.02	0
106	Vitamin A (KU / kg)	0004	26.215	3.0900	30.835	6.0519	2.9569	29	-0.76	106.02	0
106	Vitamin A (KU / kg)	2062	27.380	1.8290	30.835	6.0519	2.9569	29	-0.57	106.02	0
106	Vitamin A (KU / kg)	0027	27.469	1.4616	30.835	6.0519	2.9569	29	-0.56	106.02	0
106	Vitamin A (KU / kg)	0676	27.550	4.3000	30.835	6.0519	2.9569	29	-0.54	106.02	0
106	Vitamin A (KU / kg)	0003	27.950	0.90000	30.835	6.0519	2.9569	29	-0.48	106.02	0
106	Vitamin A (KU / kg)	0910	28.050	3.3000	30.835	6.0519	2.9569	29	-0.46	106.02	0
106	Vitamin A (KU / kg)	0098	28.200	0.20000	30.835	6.0519	2.9569	29	-0.44	106.01	0
106	Vitamin A (KU / kg)	0033	28.500	0.60000	30.835	6.0519	2.9569	29	-0.39	106.02	0
106	Vitamin A (KU / kg)	0208	28.880	1.3220	30.835	6.0519	2.9569	29	-0.32	106.02	0
106	Vitamin A (KU / kg)	0199	30.210	0.38000	30.835	6.0519	2.9569	29	-0.10	106.02	0
106	Vitamin A (KU / kg)	0227	30.850	2.1000	30.835	6.0519	2.9569	29	0.00	106.02	0
106	Vitamin A (KU / kg)	2103	30.870	2.7600	30.835	6.0519	2.9569	29	0.01	106.02	0
106	Vitamin A (KU / kg)	0861	32.300	11.400	30.835	6.0519	2.9569	29	0.24	106.02	0
106	Vitamin A (KU / kg)	0689	32.390	2.1800	30.835	6.0519	2.9569	29	0.26	106.02	0
106	Vitamin A (KU / kg)	0964	32.616	3.2209	30.835	6.0519	2.9569	29	0.29	106.00	0
106	Vitamin A (KU / kg)	0148	33.280	0.72000	30.835	6.0519	2.9569	29	0.40	106.02	0
106	Vitamin A (KU / kg)	0512	34.430	4.1000	30.835	6.0519	2.9569	29	0.59	106.02	0
106	Vitamin A (KU / kg)	0610	34.600	3.4000	30.835	6.0519	2.9569	29	0.62	106.02	0
106	Vitamin A (KU / kg)	0014	36.600	5.2000	30.835	6.0519	2.9569	29	0.95	106.02	0
106	Vitamin A (KU / kg)	0171	37.350	0.70000	30.835	6.0519	2.9569	29	1.08	106.00	0
106	Vitamin A (KU / kg)	0675	43.345	11.870	30.835	6.0519	2.9569	29	2.07	106.02	0
106	Vitamin A (KU / kg)	0038	45.100	5.4000	30.835	6.0519	2.9569	29	2.36	106.02	0
106	Vitamin A (KU / kg)	0019	51.224	4.3250	30.835	6.0519	2.9569	29	3.37	106.00	0
106	Vitamin A (KU / kg)	0169	55.000	2.0000	30.835	6.0519	2.9569	29	3.99	106.02	0
107	Vitamin B12 (µg / kg (ppb))	0227	47.200	1.4000				1		107.00	0
108	Vitamin D3 (KU / kg)	0904	2.7550	0.83000	5.3278	1.9515	0.59500	8	-1.32	108.02	0
108	Vitamin D3 (KU / kg)	0675	4.2250	0.11000	5.3278	1.9515	0.59500	8	-0.57	108.02	0
108	Vitamin D3 (KU / kg)	0610	4.4000	0.60000	5.3278	1.9515	0.59500	8	-0.48	108.02	0
108	Vitamin D3 (KU / kg)	0169	4.5000	1.0000	5.3278	1.9515	0.59500	8	-0.42	108.01	0
108	Vitamin D3 (KU / kg)	0227	5.7100	0.50000	5.3278	1.9515	0.59500	8	0.20	108.02	0
108	Vitamin D3 (KU / kg)	0861	5.7820	0.34000	5.3278	1.9515	0.59500	8	0.23	108.02	0
108	Vitamin D3 (KU / kg)	0208	7.1550	1.0300	5.3278	1.9515	0.59500	8	0.94	108.02	0
108	Vitamin D3 (KU / kg)	0676	8.0950	0.35000	5.3278	1.9515	0.59500	8	1.42	108.01	0
109	Vitamin E (IU/kg)	2068	0.04750	0.01300	118.60	31.902	6.1868	23	-3.72	109.02	0
109	Vitamin E (IU/kg)	0619	0.73800	0.00000	118.60	31.902	6.1868	23	-3.69	109.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
109	Vitamin E (IU/kg)	0027	88.646	0.53760	118.60	31.902	6.1868	23	-0.94	109.02	0
109	Vitamin E (IU/kg)	0590	97.000	2.0000	118.60	31.902	6.1868	23	-0.68	109.02	0
109	Vitamin E (IU/kg)	2062	97.100	4.8000	118.60	31.902	6.1868	23	-0.67	109.02	0
109	Vitamin E (IU/kg)	2152	99.640	0.32000	118.60	31.902	6.1868	23	-0.59	109.02	0
109	Vitamin E (IU/kg)	0169	102.50	3.0000	118.60	31.902	6.1868	23	-0.50	109.02	0
109	Vitamin E (IU/kg)	0098	104.50	21.000	118.60	31.902	6.1868	23	-0.44	109.02	0
109	Vitamin E (IU/kg)	2103	104.50	8.0000	118.60	31.902	6.1868	23	-0.44	109.02	0
109	Vitamin E (IU/kg)	0208	105.23	4.5560	118.60	31.902	6.1868	23	-0.42	109.02	0
109	Vitamin E (IU/kg)	0171	108.00	0.00000	118.60	31.902	6.1868	23	-0.33	109.99	0
109	Vitamin E (IU/kg)	0904	109.90	9.7400	118.60	31.902	6.1868	23	-0.27	109.02	0
109	Vitamin E (IU/kg)	0955	110.50	1.0000	118.60	31.902	6.1868	23	-0.25	109.02	0
109	Vitamin E (IU/kg)	0910	122.80	1.4000	118.60	31.902	6.1868	23	0.13	109.02	0
109	Vitamin E (IU/kg)	0676	128.55	1.9000	118.60	31.902	6.1868	23	0.31	109.02	0
109	Vitamin E (IU/kg)	0861	130.00	0.00000	118.60	31.902	6.1868	23	0.36	109.02	0
109	Vitamin E (IU/kg)	0610	141.50	7.0000	118.60	31.902	6.1868	23	0.72	109.02	0
109	Vitamin E (IU/kg)	0905	144.50	9.0000	118.60	31.902	6.1868	23	0.81	109.02	0
109	Vitamin E (IU/kg)	0227	147.50	41.000	118.60	31.902	6.1868	23	0.91	109.02	0
109	Vitamin E (IU/kg)	0148	155.81	9.0000	118.60	31.902	6.1868	23	1.17	109.02	0
109	Vitamin E (IU/kg)	0199	156.70	0.40000	118.60	31.902	6.1868	23	1.19	109.02	0
109	Vitamin E (IU/kg)	0563	165.01	2.9602	118.60	31.902	6.1868	23	1.45	109.02	0
109	Vitamin E (IU/kg)	0675	184.72	14.670	118.60	31.902	6.1868	23	2.07	109.02	0
112	Pyridoxine (µg / g)	0904	11.620	0.78000				2		112.01	0
112	Pyridoxine (µg / g)	0227	13.550	0.30000				2		112.01	0
113	Folic Acid (mg / kg (ppm))	0227	15.250	0.10000				2		113.01	0
113	Folic Acid (mg / kg (ppm))	0675	24.695	1.1900				2		113.02	0
114	Biotin (mg / kg (ppm))	2053	0.33500	0.03000	0.90300	0.43567	0.06100	4	-1.30	114.99	0
114	Biotin (mg / kg (ppm))	0208	0.91050	0.12900	0.90300	0.43567	0.06100	4	0.02	114.01	0
114	Biotin (mg / kg (ppm))	0227	0.97150	0.03500	0.90300	0.43567	0.06100	4	0.16	114.01	0
114	Biotin (mg / kg (ppm))	0675	1.3950	0.05000	0.90300	0.43567	0.06100	4	1.13	114.99	0
120	Alanine (%)	0148	1.0865	0.00500	1.1709	0.03604	0.01611	25	-2.34	120.05	0
120	Alanine (%)	0626	1.0920	0.03200	1.1709	0.03604	0.01611	25	-2.19	120.05	0
120	Alanine (%)	0684	1.1090	0.03200	1.1709	0.03604	0.01611	25	-1.72	120.00	0
120	Alanine (%)	0610	1.1150	0.01000	1.1709	0.03604	0.01611	25	-1.55	120.05	0
120	Alanine (%)	0652	1.1300	0.02000	1.1709	0.03604	0.01611	25	-1.14	120.00	0
120	Alanine (%)	0676	1.1450	0.01000	1.1709	0.03604	0.01611	25	-0.72	120.05	0
120	Alanine (%)	0619	1.1600	0.04000	1.1709	0.03604	0.01611	25	-0.30	120.00	0
120	Alanine (%)	0968	1.1665	0.01100	1.1709	0.03604	0.01611	25	-0.12	120.00	0
120	Alanine (%)	2059	1.1715	0.00300	1.1709	0.03604	0.01611	25	0.02	120.00	0
120	Alanine (%)	0504	1.1750	0.03000	1.1709	0.03604	0.01611	25	0.11	120.00	0
120	Alanine (%)	0675	1.1750	0.01000	1.1709	0.03604	0.01611	25	0.11	120.00	0
120	Alanine (%)	0941	1.1750	0.03000	1.1709	0.03604	0.01611	25	0.11	120.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
120	Alanine (%)	0904	1.1750	0.05000	1.1709	0.03604	0.01611	25	0.11	120.99	0
120	Alanine (%)	0644	1.1760	0.01200	1.1709	0.03604	0.01611	25	0.14	120.00	0
120	Alanine (%)	0571	1.1790	0.01400	1.1709	0.03604	0.01611	25	0.22	120.00	0
120	Alanine (%)	0859	1.1795	0.00700	1.1709	0.03604	0.01611	25	0.24	120.00	0
120	Alanine (%)	0171	1.1800	0.00000	1.1709	0.03604	0.01611	25	0.25	120.00	0
120	Alanine (%)	0872	1.1830	0.00600	1.1709	0.03604	0.01611	25	0.33	120.00	0
120	Alanine (%)	0910	1.1850	0.01000	1.1709	0.03604	0.01611	25	0.39	120.00	0
120	Alanine (%)	0939	1.1900	0.00000	1.1709	0.03604	0.01611	25	0.53	120.05	0
120	Alanine (%)	0227	1.2000	0.00000	1.1709	0.03604	0.01611	25	0.81	120.00	0
120	Alanine (%)	0098	1.2045	0.01900	1.1709	0.03604	0.01611	25	0.93	120.02	0
120	Alanine (%)	0354	1.2060	0.02400	1.1709	0.03604	0.01611	25	0.97	120.00	0
120	Alanine (%)	0918	1.2331	0.02770	1.1709	0.03604	0.01611	25	1.72	120.00	0
120	Alanine (%)	0226	2.6000	0.00000	1.1709	0.03604	0.01611	25	39.65	120.99	0
121	Arginine (%)	0148	0.83500	0.00400	0.95311	0.02859	0.01774	25	-4.13	121.05	0
121	Arginine (%)	0684	0.89050	0.00900	0.95311	0.02859	0.01774	25	-2.19	121.00	0
121	Arginine (%)	0171	0.92500	0.01000	0.95311	0.02859	0.01774	25	-0.98	121.00	0
121	Arginine (%)	0626	0.92700	0.01800	0.95311	0.02859	0.01774	25	-0.91	121.05	0
121	Arginine (%)	0652	0.93000	0.00000	0.95311	0.02859	0.01774	25	-0.81	121.00	0
121	Arginine (%)	0227	0.93500	0.01000	0.95311	0.02859	0.01774	25	-0.63	121.00	0
121	Arginine (%)	0504	0.93500	0.03000	0.95311	0.02859	0.01774	25	-0.63	121.00	0
121	Arginine (%)	0675	0.93500	0.01000	0.95311	0.02859	0.01774	25	-0.63	121.00	0
121	Arginine (%)	0859	0.93950	0.00300	0.95311	0.02859	0.01774	25	-0.48	121.00	0
121	Arginine (%)	0939	0.94000	0.00000	0.95311	0.02859	0.01774	25	-0.46	121.05	0
121	Arginine (%)	0098	0.94150	0.03100	0.95311	0.02859	0.01774	25	-0.41	121.02	0
121	Arginine (%)	0918	0.94490	0.00340	0.95311	0.02859	0.01774	25	-0.29	121.00	0
121	Arginine (%)	0872	0.95000	0.00400	0.95311	0.02859	0.01774	25	-0.11	121.00	0
121	Arginine (%)	0941	0.95500	0.01000	0.95311	0.02859	0.01774	25	0.07	121.00	0
121	Arginine (%)	2059	0.96200	0.00400	0.95311	0.02859	0.01774	25	0.31	121.00	0
121	Arginine (%)	0968	0.96500	0.00200	0.95311	0.02859	0.01774	25	0.42	121.00	0
121	Arginine (%)	0610	0.96500	0.01000	0.95311	0.02859	0.01774	25	0.42	121.05	0
121	Arginine (%)	0354	0.96700	0.01400	0.95311	0.02859	0.01774	25	0.49	121.00	0
121	Arginine (%)	0676	0.97000	0.02000	0.95311	0.02859	0.01774	25	0.59	121.05	0
121	Arginine (%)	0644	0.97350	0.01900	0.95311	0.02859	0.01774	25	0.71	121.00	0
121	Arginine (%)	0904	0.98000	0.00000	0.95311	0.02859	0.01774	25	0.94	121.99	0
121	Arginine (%)	0571	0.98600	0.01000	0.95311	0.02859	0.01774	25	1.15	121.00	0
121	Arginine (%)	0619	0.98900	0.00200	0.95311	0.02859	0.01774	25	1.26	121.00	0
121	Arginine (%)	0910	1.0450	0.01000	0.95311	0.02859	0.01774	25	3.21	121.00	0
121	Arginine (%)	0226	2.7550	0.21000	0.95311	0.02859	0.01774	25	63.02	121.99	0
122	Aspartic (%)	0904	0.86500	0.03000	1.3017	0.03380	0.02008	25	-12.92	122.99	0
122	Aspartic (%)	0652	1.2550	0.03000	1.3017	0.03380	0.02008	25	-1.38	122.00	0
122	Aspartic (%)	0626	1.2640	0.03600	1.3017	0.03380	0.02008	25	-1.12	122.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
122	Aspartic (%)	0684	1.2740	0.01000	1.3017	0.03380	0.02008	25	-0.82	122.00	0
122	Aspartic (%)	0872	1.2745	0.01100	1.3017	0.03380	0.02008	25	-0.81	122.00	0
122	Aspartic (%)	0675	1.2750	0.01000	1.3017	0.03380	0.02008	25	-0.79	122.00	0
122	Aspartic (%)	0941	1.2850	0.05000	1.3017	0.03380	0.02008	25	-0.50	122.00	0
122	Aspartic (%)	0610	1.2850	0.01000	1.3017	0.03380	0.02008	25	-0.50	122.05	0
122	Aspartic (%)	0859	1.2875	0.00100	1.3017	0.03380	0.02008	25	-0.42	122.00	0
122	Aspartic (%)	0939	1.2900	0.00000	1.3017	0.03380	0.02008	25	-0.35	122.05	0
122	Aspartic (%)	2059	1.2950	0.00000	1.3017	0.03380	0.02008	25	-0.20	122.00	0
122	Aspartic (%)	0644	1.2965	0.01300	1.3017	0.03380	0.02008	25	-0.16	122.00	0
122	Aspartic (%)	0098	1.2975	0.06700	1.3017	0.03380	0.02008	25	-0.13	122.02	0
122	Aspartic (%)	0968	1.3020	0.01600	1.3017	0.03380	0.02008	25	0.01	122.00	0
122	Aspartic (%)	0148	1.3050	0.00000	1.3017	0.03380	0.02008	25	0.10	122.05	0
122	Aspartic (%)	0504	1.3050	0.03000	1.3017	0.03380	0.02008	25	0.10	122.00	0
122	Aspartic (%)	0571	1.3070	0.00200	1.3017	0.03380	0.02008	25	0.16	122.00	0
122	Aspartic (%)	0227	1.3150	0.01000	1.3017	0.03380	0.02008	25	0.39	122.00	0
122	Aspartic (%)	0676	1.3150	0.01000	1.3017	0.03380	0.02008	25	0.39	122.05	0
122	Aspartic (%)	0171	1.3200	0.02000	1.3017	0.03380	0.02008	25	0.54	122.00	0
122	Aspartic (%)	0354	1.3450	0.01400	1.3017	0.03380	0.02008	25	1.28	122.00	0
122	Aspartic (%)	0619	1.3450	0.03000	1.3017	0.03380	0.02008	25	1.28	122.00	0
122	Aspartic (%)	0910	1.3500	0.00000	1.3017	0.03380	0.02008	25	1.43	122.00	0
122	Aspartic (%)	0918	1.3523	0.03190	1.3017	0.03380	0.02008	25	1.49	122.00	0
122	Aspartic (%)	0226	3.1150	0.07000	1.3017	0.03380	0.02008	25	53.64	122.99	0
124	Cysteine/Cystine (%)	0148	0.28750	0.00700	0.36749	0.02837	0.00829	23	-2.82	124.05	0
124	Cysteine/Cystine (%)	0227	0.32000	0.00000	0.36749	0.02837	0.00829	23	-1.67	124.00	0
124	Cysteine/Cystine (%)	0652	0.33500	0.01000	0.36749	0.02837	0.00829	23	-1.15	124.00	0
124	Cysteine/Cystine (%)	0644	0.33650	0.00100	0.36749	0.02837	0.00829	23	-1.09	124.00	0
124	Cysteine/Cystine (%)	0918	0.34525	0.02170	0.36749	0.02837	0.00829	23	-0.78	124.00	0
124	Cysteine/Cystine (%)	0504	0.35500	0.01000	0.36749	0.02837	0.00829	23	-0.44	124.00	0
124	Cysteine/Cystine (%)	0684	0.36000	0.00400	0.36749	0.02837	0.00829	23	-0.26	124.00	0
124	Cysteine/Cystine (%)	0968	0.36200	0.00800	0.36749	0.02837	0.00829	23	-0.19	124.00	0
124	Cysteine/Cystine (%)	0571	0.36300	0.00200	0.36749	0.02837	0.00829	23	-0.16	124.00	0
124	Cysteine/Cystine (%)	0910	0.36500	0.01000	0.36749	0.02837	0.00829	23	-0.09	124.00	0
124	Cysteine/Cystine (%)	0098	0.36500	0.00200	0.36749	0.02837	0.00829	23	-0.09	124.02	0
124	Cysteine/Cystine (%)	0939	0.36500	0.01000	0.36749	0.02837	0.00829	23	-0.09	124.05	0
124	Cysteine/Cystine (%)	0872	0.36550	0.00900	0.36749	0.02837	0.00829	23	-0.07	124.00	0
124	Cysteine/Cystine (%)	2059	0.36900	0.00400	0.36749	0.02837	0.00829	23	0.05	124.00	0
124	Cysteine/Cystine (%)	0354	0.37100	0.04600	0.36749	0.02837	0.00829	23	0.12	124.00	0
124	Cysteine/Cystine (%)	0619	0.37250	0.01300	0.36749	0.02837	0.00829	23	0.18	124.00	0
124	Cysteine/Cystine (%)	0941	0.37300	0.00000	0.36749	0.02837	0.00829	23	0.19	124.00	0
124	Cysteine/Cystine (%)	0610	0.38000	0.00000	0.36749	0.02837	0.00829	23	0.44	124.05	0
124	Cysteine/Cystine (%)	0859	0.38450	0.00300	0.36749	0.02837	0.00829	23	0.60	124.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
124	Cysteine/Cystine (%)	0171	0.40500	0.01000	0.36749	0.02837	0.00829	23	1.32	124.00	0
124	Cysteine/Cystine (%)	0675	0.41000	0.02000	0.36749	0.02837	0.00829	23	1.50	124.00	0
124	Cysteine/Cystine (%)	0676	0.42000	0.00000	0.36749	0.02837	0.00829	23	1.85	124.05	0
124	Cysteine/Cystine (%)	0226	0.55000	0.00000	0.36749	0.02837	0.00829	23	6.43	124.99	0
125	Glutamic (%)	0904	2.9800	0.10000	3.3941	0.10190	0.05115	25	-4.06	125.99	0
125	Glutamic (%)	0684	3.2180	0.10400	3.3941	0.10190	0.05115	25	-1.73	125.00	0
125	Glutamic (%)	0626	3.2425	0.03100	3.3941	0.10190	0.05115	25	-1.49	125.05	0
125	Glutamic (%)	0676	3.2700	0.04000	3.3941	0.10190	0.05115	25	-1.22	125.05	0
125	Glutamic (%)	0941	3.3400	0.06000	3.3941	0.10190	0.05115	25	-0.53	125.00	0
125	Glutamic (%)	0148	3.3420	0.02400	3.3941	0.10190	0.05115	25	-0.51	125.05	0
125	Glutamic (%)	2059	3.3465	0.00500	3.3941	0.10190	0.05115	25	-0.47	125.00	0
125	Glutamic (%)	0968	3.3515	0.00900	3.3941	0.10190	0.05115	25	-0.42	125.00	0
125	Glutamic (%)	0859	3.3600	0.00800	3.3941	0.10190	0.05115	25	-0.33	125.00	0
125	Glutamic (%)	0504	3.3650	0.07000	3.3941	0.10190	0.05115	25	-0.29	125.00	0
125	Glutamic (%)	0610	3.3650	0.03000	3.3941	0.10190	0.05115	25	-0.29	125.05	0
125	Glutamic (%)	0571	3.3750	0.00600	3.3941	0.10190	0.05115	25	-0.19	125.00	0
125	Glutamic (%)	0098	3.3800	0.11000	3.3941	0.10190	0.05115	25	-0.14	125.02	0
125	Glutamic (%)	0644	3.4070	0.02600	3.3941	0.10190	0.05115	25	0.13	125.00	0
125	Glutamic (%)	0619	3.4100	0.06000	3.3941	0.10190	0.05115	25	0.16	125.00	0
125	Glutamic (%)	0227	3.4150	0.01000	3.3941	0.10190	0.05115	25	0.21	125.00	0
125	Glutamic (%)	0652	3.4400	0.08000	3.3941	0.10190	0.05115	25	0.45	125.00	0
125	Glutamic (%)	0354	3.4425	0.08100	3.3941	0.10190	0.05115	25	0.48	125.00	0
125	Glutamic (%)	0675	3.4450	0.01000	3.3941	0.10190	0.05115	25	0.50	125.00	0
125	Glutamic (%)	0918	3.4712	0.06980	3.3941	0.10190	0.05115	25	0.76	125.00	0
125	Glutamic (%)	0872	3.4895	0.00500	3.3941	0.10190	0.05115	25	0.94	125.00	0
125	Glutamic (%)	0171	3.5050	0.01000	3.3941	0.10190	0.05115	25	1.09	125.00	0
125	Glutamic (%)	0939	3.5300	0.00000	3.3941	0.10190	0.05115	25	1.33	125.05	0
125	Glutamic (%)	0910	3.5300	0.02000	3.3941	0.10190	0.05115	25	1.33	125.00	0
125	Glutamic (%)	0226	8.1050	0.31000	3.3941	0.10190	0.05115	25	46.23	125.99	0
126	Glycine (%)	0171	0.71500	0.01000	0.77849	0.01760	0.01039	25	-3.61	126.00	0
126	Glycine (%)	0652	0.74500	0.01000	0.77849	0.01760	0.01039	25	-1.90	126.00	0
126	Glycine (%)	0939	0.74500	0.01000	0.77849	0.01760	0.01039	25	-1.90	126.05	0
126	Glycine (%)	0148	0.75450	0.00500	0.77849	0.01760	0.01039	25	-1.36	126.05	0
126	Glycine (%)	0098	0.76300	0.01000	0.77849	0.01760	0.01039	25	-0.88	126.02	0
126	Glycine (%)	0626	0.76900	0.02200	0.77849	0.01760	0.01039	25	-0.54	126.05	0
126	Glycine (%)	0675	0.77000	0.00000	0.77849	0.01760	0.01039	25	-0.48	126.00	0
126	Glycine (%)	0910	0.77000	0.02000	0.77849	0.01760	0.01039	25	-0.48	126.00	0
126	Glycine (%)	0968	0.77500	0.00000	0.77849	0.01760	0.01039	25	-0.20	126.00	0
126	Glycine (%)	0859	0.77550	0.00100	0.77849	0.01760	0.01039	25	-0.17	126.00	0
126	Glycine (%)	0918	0.77650	0.00480	0.77849	0.01760	0.01039	25	-0.11	126.00	0
126	Glycine (%)	0684	0.77900	0.01200	0.77849	0.01760	0.01039	25	0.03	126.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
126	Glycine (%)	0941	0.78000	0.02000	0.77849	0.01760	0.01039	25	0.09	126.00	0
126	Glycine (%)	0610	0.78000	0.00000	0.77849	0.01760	0.01039	25	0.09	126.05	0
126	Glycine (%)	0619	0.78050	0.01300	0.77849	0.01760	0.01039	25	0.11	126.00	0
126	Glycine (%)	2059	0.78050	0.00300	0.77849	0.01760	0.01039	25	0.11	126.00	0
126	Glycine (%)	0504	0.79000	0.02000	0.77849	0.01760	0.01039	25	0.65	126.00	0
126	Glycine (%)	0676	0.79000	0.02000	0.77849	0.01760	0.01039	25	0.65	126.05	0
126	Glycine (%)	0354	0.79050	0.00700	0.77849	0.01760	0.01039	25	0.68	126.00	0
126	Glycine (%)	0644	0.79050	0.01100	0.77849	0.01760	0.01039	25	0.68	126.00	0
126	Glycine (%)	0571	0.79250	0.00500	0.77849	0.01760	0.01039	25	0.80	126.00	0
126	Glycine (%)	0227	0.79500	0.01000	0.77849	0.01760	0.01039	25	0.94	126.00	0
126	Glycine (%)	0872	0.79900	0.00600	0.77849	0.01760	0.01039	25	1.17	126.00	0
126	Glycine (%)	0904	0.80000	0.04000	0.77849	0.01760	0.01039	25	1.22	126.99	0
126	Glycine (%)	0226	1.5100	0.00000	0.77849	0.01760	0.01039	25	41.56	126.99	0
127	Histidine (%)	0148	0.39750	0.01100	0.46515	0.02184	0.01404	25	-3.10	127.05	0
127	Histidine (%)	0859	0.42300	0.00200	0.46515	0.02184	0.01404	25	-1.93	127.00	0
127	Histidine (%)	0939	0.43500	0.01000	0.46515	0.02184	0.01404	25	-1.38	127.05	0
127	Histidine (%)	0684	0.44100	0.01400	0.46515	0.02184	0.01404	25	-1.11	127.00	0
127	Histidine (%)	0675	0.44500	0.01000	0.46515	0.02184	0.01404	25	-0.92	127.00	0
127	Histidine (%)	0626	0.45150	0.01100	0.46515	0.02184	0.01404	25	-0.62	127.05	0
127	Histidine (%)	2059	0.45450	0.00100	0.46515	0.02184	0.01404	25	-0.49	127.00	0
127	Histidine (%)	0098	0.45650	0.02500	0.46515	0.02184	0.01404	25	-0.40	127.02	0
127	Histidine (%)	0918	0.45865	0.00090	0.46515	0.02184	0.01404	25	-0.30	127.00	0
127	Histidine (%)	0968	0.46300	0.00600	0.46515	0.02184	0.01404	25	-0.10	127.00	0
127	Histidine (%)	0171	0.46500	0.01000	0.46515	0.02184	0.01404	25	-0.01	127.00	0
127	Histidine (%)	0910	0.46500	0.01000	0.46515	0.02184	0.01404	25	-0.01	127.00	0
127	Histidine (%)	0941	0.46500	0.01000	0.46515	0.02184	0.01404	25	-0.01	127.00	0
127	Histidine (%)	0872	0.46850	0.00100	0.46515	0.02184	0.01404	25	0.15	127.00	0
127	Histidine (%)	0652	0.47000	0.00000	0.46515	0.02184	0.01404	25	0.22	127.00	0
127	Histidine (%)	0610	0.47000	0.00000	0.46515	0.02184	0.01404	25	0.22	127.05	0
127	Histidine (%)	0676	0.47000	0.00000	0.46515	0.02184	0.01404	25	0.22	127.05	0
127	Histidine (%)	0571	0.47400	0.02000	0.46515	0.02184	0.01404	25	0.41	127.00	0
127	Histidine (%)	0354	0.47450	0.00300	0.46515	0.02184	0.01404	25	0.43	127.00	0
127	Histidine (%)	0227	0.47500	0.01000	0.46515	0.02184	0.01404	25	0.45	127.00	0
127	Histidine (%)	0504	0.47500	0.01000	0.46515	0.02184	0.01404	25	0.45	127.00	0
127	Histidine (%)	0619	0.49300	0.01200	0.46515	0.02184	0.01404	25	1.28	127.00	0
127	Histidine (%)	0644	0.50200	0.00400	0.46515	0.02184	0.01404	25	1.69	127.00	0
127	Histidine (%)	0904	0.51500	0.01000	0.46515	0.02184	0.01404	25	2.28	127.99	0
127	Histidine (%)	0226	1.0100	0.16000	0.46515	0.02184	0.01404	25	24.95	127.99	0
128	Isoleucine (%)	0148	0.52600	0.01000	0.67723	0.03686	0.01362	25	-4.10	128.05	0
128	Isoleucine (%)	0684	0.58800	0.02400	0.67723	0.03686	0.01362	25	-2.42	128.00	0
128	Isoleucine (%)	0939	0.61500	0.01000	0.67723	0.03686	0.01362	25	-1.69	128.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
128	Isoleucine (%)	0171	0.63500	0.01000	0.67723	0.03686	0.01362	25	-1.15	128.00	0
128	Isoleucine (%)	0918	0.64345	0.01550	0.67723	0.03686	0.01362	25	-0.92	128.00	0
128	Isoleucine (%)	0859	0.64800	0.01200	0.67723	0.03686	0.01362	25	-0.79	128.00	0
128	Isoleucine (%)	0675	0.65500	0.01000	0.67723	0.03686	0.01362	25	-0.60	128.00	0
128	Isoleucine (%)	0968	0.66100	0.01200	0.67723	0.03686	0.01362	25	-0.44	128.00	0
128	Isoleucine (%)	0652	0.66500	0.01000	0.67723	0.03686	0.01362	25	-0.33	128.00	0
128	Isoleucine (%)	0872	0.67100	0.01000	0.67723	0.03686	0.01362	25	-0.17	128.00	0
128	Isoleucine (%)	0626	0.67550	0.01500	0.67723	0.03686	0.01362	25	-0.05	128.05	0
128	Isoleucine (%)	0571	0.68300	0.00600	0.67723	0.03686	0.01362	25	0.16	128.00	0
128	Isoleucine (%)	2059	0.68400	0.00200	0.67723	0.03686	0.01362	25	0.18	128.00	0
128	Isoleucine (%)	0941	0.68500	0.01000	0.67723	0.03686	0.01362	25	0.21	128.00	0
128	Isoleucine (%)	0098	0.68500	0.00600	0.67723	0.03686	0.01362	25	0.21	128.02	0
128	Isoleucine (%)	0619	0.68850	0.01500	0.67723	0.03686	0.01362	25	0.31	128.00	0
128	Isoleucine (%)	0354	0.69150	0.00100	0.67723	0.03686	0.01362	25	0.39	128.00	0
128	Isoleucine (%)	0610	0.69500	0.01000	0.67723	0.03686	0.01362	25	0.48	128.05	0
128	Isoleucine (%)	0227	0.70000	0.02000	0.67723	0.03686	0.01362	25	0.62	128.00	0
128	Isoleucine (%)	0910	0.70500	0.03000	0.67723	0.03686	0.01362	25	0.75	128.00	0
128	Isoleucine (%)	0676	0.70500	0.01000	0.67723	0.03686	0.01362	25	0.75	128.05	0
128	Isoleucine (%)	0644	0.70900	0.00200	0.67723	0.03686	0.01362	25	0.86	128.00	0
128	Isoleucine (%)	0904	0.71500	0.07000	0.67723	0.03686	0.01362	25	1.02	128.99	0
128	Isoleucine (%)	0504	0.74000	0.02000	0.67723	0.03686	0.01362	25	1.70	128.00	0
128	Isoleucine (%)	0226	1.5300	0.00000	0.67723	0.03686	0.01362	25	23.14	128.99	0
129	Leucine (%)	0148	1.7670	0.03000	1.8869	0.04862	0.03238	25	-2.47	129.05	0
129	Leucine (%)	0684	1.7715	0.06100	1.8869	0.04862	0.03238	25	-2.37	129.00	0
129	Leucine (%)	0968	1.8260	0.08200	1.8869	0.04862	0.03238	25	-1.25	129.00	0
129	Leucine (%)	0939	1.8300	0.00000	1.8869	0.04862	0.03238	25	-1.17	129.05	0
129	Leucine (%)	0652	1.8450	0.03000	1.8869	0.04862	0.03238	25	-0.86	129.00	0
129	Leucine (%)	2059	1.8525	0.01500	1.8869	0.04862	0.03238	25	-0.71	129.00	0
129	Leucine (%)	0626	1.8650	0.04200	1.8869	0.04862	0.03238	25	-0.45	129.05	0
129	Leucine (%)	0171	1.8650	0.01000	1.8869	0.04862	0.03238	25	-0.45	129.00	0
129	Leucine (%)	0676	1.8700	0.02000	1.8869	0.04862	0.03238	25	-0.35	129.05	0
129	Leucine (%)	0859	1.8790	0.02200	1.8869	0.04862	0.03238	25	-0.16	129.00	0
129	Leucine (%)	0675	1.8800	0.02000	1.8869	0.04862	0.03238	25	-0.14	129.00	0
129	Leucine (%)	0941	1.8850	0.03000	1.8869	0.04862	0.03238	25	-0.04	129.00	0
129	Leucine (%)	0227	1.8900	0.02000	1.8869	0.04862	0.03238	25	0.06	129.00	0
129	Leucine (%)	0644	1.8910	0.00200	1.8869	0.04862	0.03238	25	0.08	129.00	0
129	Leucine (%)	0872	1.8950	0.00800	1.8869	0.04862	0.03238	25	0.17	129.00	0
129	Leucine (%)	0904	1.8950	0.01000	1.8869	0.04862	0.03238	25	0.17	129.99	0
129	Leucine (%)	0619	1.9000	0.00000	1.8869	0.04862	0.03238	25	0.27	129.00	0
129	Leucine (%)	0098	1.9075	0.02100	1.8869	0.04862	0.03238	25	0.42	129.02	0
129	Leucine (%)	0571	1.9095	0.00500	1.8869	0.04862	0.03238	25	0.46	129.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
129	Leucine (%)	0910	1.9200	0.06000	1.8869	0.04862	0.03238	25	0.68	129.00	0
129	Leucine (%)	0610	1.9200	0.02000	1.8869	0.04862	0.03238	25	0.68	129.05	0
129	Leucine (%)	0504	1.9400	0.06000	1.8869	0.04862	0.03238	25	1.09	129.00	0
129	Leucine (%)	0354	1.9635	0.03100	1.8869	0.04862	0.03238	25	1.58	129.00	0
129	Leucine (%)	0918	1.9782	0.04040	1.8869	0.04862	0.03238	25	1.88	129.00	0
129	Leucine (%)	0226	4.0550	0.17000	1.8869	0.04862	0.03238	25	44.60	129.99	0
130	L-Lysine (%)	0148	0.77600	0.03400	0.85368	0.05159	0.01769	30	-1.51	130.05	0
130	L-Lysine (%)	0684	0.77800	0.00400	0.85368	0.05159	0.01769	30	-1.47	130.00	0
130	L-Lysine (%)	0859	0.79050	0.00100	0.85368	0.05159	0.01769	30	-1.22	130.00	0
130	L-Lysine (%)	0626	0.79450	0.00700	0.85368	0.05159	0.01769	30	-1.15	130.05	0
130	L-Lysine (%)	0652	0.80500	0.01000	0.85368	0.05159	0.01769	30	-0.94	130.00	0
130	L-Lysine (%)	0939	0.80500	0.01000	0.85368	0.05159	0.01769	30	-0.94	130.05	0
130	L-Lysine (%)	0512	0.80805	0.00790	0.85368	0.05159	0.01769	30	-0.88	130.00	0
130	L-Lysine (%)	0941	0.82500	0.01000	0.85368	0.05159	0.01769	30	-0.56	130.00	0
130	L-Lysine (%)	0872	0.83300	0.01600	0.85368	0.05159	0.01769	30	-0.40	130.00	0
130	L-Lysine (%)	2059	0.83600	0.00000	0.85368	0.05159	0.01769	30	-0.34	130.00	0
130	L-Lysine (%)	0016	0.83700	0.02600	0.85368	0.05159	0.01769	30	-0.32	130.05	0
130	L-Lysine (%)	0918	0.83795	0.00290	0.85368	0.05159	0.01769	30	-0.30	130.00	0
130	L-Lysine (%)	0610	0.84000	0.02000	0.85368	0.05159	0.01769	30	-0.27	130.05	0
130	L-Lysine (%)	0675	0.84500	0.01000	0.85368	0.05159	0.01769	30	-0.17	130.00	0
130	L-Lysine (%)	0644	0.84800	0.01800	0.85368	0.05159	0.01769	30	-0.11	130.00	0
130	L-Lysine (%)	0571	0.84850	0.00500	0.85368	0.05159	0.01769	30	-0.10	130.00	0
130	L-Lysine (%)	0619	0.85250	0.03100	0.85368	0.05159	0.01769	30	-0.02	130.00	0
130	L-Lysine (%)	0968	0.85500	0.01000	0.85368	0.05159	0.01769	30	0.03	130.00	0
130	L-Lysine (%)	0676	0.85500	0.03000	0.85368	0.05159	0.01769	30	0.03	130.05	0
130	L-Lysine (%)	0354	0.88150	0.01700	0.85368	0.05159	0.01769	30	0.54	130.00	0
130	L-Lysine (%)	0014	0.88300	0.00800	0.85368	0.05159	0.01769	30	0.57	130.05	0
130	L-Lysine (%)	0910	0.88500	0.07000	0.85368	0.05159	0.01769	30	0.61	130.00	0
130	L-Lysine (%)	0904	0.88500	0.01000	0.85368	0.05159	0.01769	30	0.61	130.99	0
130	L-Lysine (%)	0504	0.89500	0.01000	0.85368	0.05159	0.01769	30	0.80	130.00	0
130	L-Lysine (%)	0027	0.89800	0.00400	0.85368	0.05159	0.01769	30	0.86	130.05	0
130	L-Lysine (%)	0098	0.90450	0.02900	0.85368	0.05159	0.01769	30	0.98	130.02	0
130	L-Lysine (%)	0848	0.91500	0.03000	0.85368	0.05159	0.01769	30	1.19	130.00	0
130	L-Lysine (%)	0171	0.95000	0.00000	0.85368	0.05159	0.01769	30	1.87	130.00	0
130	L-Lysine (%)	0227	0.98500	0.05000	0.85368	0.05159	0.01769	30	2.55	130.00	0
130	L-Lysine (%)	0226	1.6750	0.05000	0.85368	0.05159	0.01769	30	15.92	130.99	0
131	Methionine (%)	0148	0.32000	0.00200	0.40796	0.03712	0.01943	28	-2.37	131.05	0
131	Methionine (%)	0626	0.34650	0.00900	0.40796	0.03712	0.01943	28	-1.66	131.05	0
131	Methionine (%)	0939	0.37000	0.02000	0.40796	0.03712	0.01943	28	-1.02	131.05	0
131	Methionine (%)	0512	0.37130	0.00380	0.40796	0.03712	0.01943	28	-0.99	131.00	0
131	Methionine (%)	0027	0.37750	0.05700	0.40796	0.03712	0.01943	28	-0.82	131.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
131	Methionine (%)	0098	0.38150	0.00500	0.40796	0.03712	0.01943	28	-0.71	131.02	0
131	Methionine (%)	0941	0.38900	0.00000	0.40796	0.03712	0.01943	28	-0.51	131.00	0
131	Methionine (%)	0918	0.39095	0.01030	0.40796	0.03712	0.01943	28	-0.46	131.00	0
131	Methionine (%)	0859	0.39250	0.00100	0.40796	0.03712	0.01943	28	-0.42	131.00	0
131	Methionine (%)	0619	0.39500	0.00000	0.40796	0.03712	0.01943	28	-0.35	131.00	0
131	Methionine (%)	2059	0.39900	0.00200	0.40796	0.03712	0.01943	28	-0.24	131.00	0
131	Methionine (%)	0227	0.40000	0.00000	0.40796	0.03712	0.01943	28	-0.21	131.00	0
131	Methionine (%)	0014	0.40150	0.13500	0.40796	0.03712	0.01943	28	-0.17	131.05	0
131	Methionine (%)	0571	0.40150	0.00700	0.40796	0.03712	0.01943	28	-0.17	131.00	0
131	Methionine (%)	0872	0.40500	0.00600	0.40796	0.03712	0.01943	28	-0.08	131.00	0
131	Methionine (%)	0610	0.40500	0.01000	0.40796	0.03712	0.01943	28	-0.08	131.05	0
131	Methionine (%)	0354	0.40800	0.04600	0.40796	0.03712	0.01943	28	0.00	131.00	0
131	Methionine (%)	0504	0.41000	0.00000	0.40796	0.03712	0.01943	28	0.05	131.00	0
131	Methionine (%)	0652	0.41000	0.02000	0.40796	0.03712	0.01943	28	0.05	131.00	0
131	Methionine (%)	0904	0.41000	0.04000	0.40796	0.03712	0.01943	28	0.05	131.99	0
131	Methionine (%)	0968	0.41850	0.00500	0.40796	0.03712	0.01943	28	0.28	131.00	0
131	Methionine (%)	0910	0.42000	0.00000	0.40796	0.03712	0.01943	28	0.32	131.00	0
131	Methionine (%)	0848	0.44500	0.01000	0.40796	0.03712	0.01943	28	1.00	131.00	0
131	Methionine (%)	0644	0.46250	0.04300	0.40796	0.03712	0.01943	28	1.47	131.00	0
131	Methionine (%)	0171	0.47000	0.00000	0.40796	0.03712	0.01943	28	1.67	131.00	0
131	Methionine (%)	0684	0.47100	0.00200	0.40796	0.03712	0.01943	28	1.70	131.00	0
131	Methionine (%)	0675	0.50000	0.02000	0.40796	0.03712	0.01943	28	2.48	131.00	0
131	Methionine (%)	0226	0.71500	0.09000	0.40796	0.03712	0.01943	28	8.27	131.99	0
132	Phenylalanine (%)	0148	0.80750	0.03100	0.89301	0.03653	0.01844	25	-2.34	132.05	0
132	Phenylalanine (%)	0684	0.81250	0.01700	0.89301	0.03653	0.01844	25	-2.20	132.00	0
132	Phenylalanine (%)	0652	0.86000	0.02000	0.89301	0.03653	0.01844	25	-0.90	132.00	0
132	Phenylalanine (%)	0644	0.86250	0.00100	0.89301	0.03653	0.01844	25	-0.84	132.00	0
132	Phenylalanine (%)	2059	0.86750	0.00500	0.89301	0.03653	0.01844	25	-0.70	132.00	0
132	Phenylalanine (%)	0626	0.86800	0.02000	0.89301	0.03653	0.01844	25	-0.68	132.05	0
132	Phenylalanine (%)	0941	0.87000	0.00000	0.89301	0.03653	0.01844	25	-0.63	132.00	0
132	Phenylalanine (%)	0098	0.88000	0.02600	0.89301	0.03653	0.01844	25	-0.36	132.02	0
132	Phenylalanine (%)	0904	0.88000	0.08000	0.89301	0.03653	0.01844	25	-0.36	132.99	0
132	Phenylalanine (%)	0968	0.88050	0.00500	0.89301	0.03653	0.01844	25	-0.34	132.00	0
132	Phenylalanine (%)	0227	0.88500	0.01000	0.89301	0.03653	0.01844	25	-0.22	132.00	0
132	Phenylalanine (%)	0504	0.88500	0.03000	0.89301	0.03653	0.01844	25	-0.22	132.00	0
132	Phenylalanine (%)	0675	0.88500	0.01000	0.89301	0.03653	0.01844	25	-0.22	132.00	0
132	Phenylalanine (%)	0676	0.88500	0.01000	0.89301	0.03653	0.01844	25	-0.22	132.05	0
132	Phenylalanine (%)	0571	0.88650	0.01100	0.89301	0.03653	0.01844	25	-0.18	132.00	0
132	Phenylalanine (%)	0918	0.89335	0.01510	0.89301	0.03653	0.01844	25	0.01	132.00	0
132	Phenylalanine (%)	0859	0.89850	0.01500	0.89301	0.03653	0.01844	25	0.15	132.00	0
132	Phenylalanine (%)	0171	0.91000	0.00000	0.89301	0.03653	0.01844	25	0.47	132.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
132	Phenylalanine (%)	0872	0.91300	0.00800	0.89301	0.03653	0.01844	25	0.55	132.00	0
132	Phenylalanine (%)	0619	0.92900	0.02400	0.89301	0.03653	0.01844	25	0.99	132.00	0
132	Phenylalanine (%)	0610	0.93000	0.00000	0.89301	0.03653	0.01844	25	1.01	132.05	0
132	Phenylalanine (%)	0354	0.93650	0.03300	0.89301	0.03653	0.01844	25	1.19	132.00	0
132	Phenylalanine (%)	0939	0.95500	0.01000	0.89301	0.03653	0.01844	25	1.70	132.05	0
132	Phenylalanine (%)	0910	1.0650	0.05000	0.89301	0.03653	0.01844	25	4.71	132.00	0
132	Phenylalanine (%)	0226	1.8650	0.03000	0.89301	0.03653	0.01844	25	26.61	132.99	0
133	Proline (%)	0684	1.2850	0.05200	1.3756	0.06071	0.02685	24	-1.49	133.00	0
133	Proline (%)	0872	1.2940	0.00200	1.3756	0.06071	0.02685	24	-1.34	133.00	0
133	Proline (%)	0171	1.3100	0.00000	1.3756	0.06071	0.02685	24	-1.08	133.00	0
133	Proline (%)	0148	1.3165	0.02500	1.3756	0.06071	0.02685	24	-0.97	133.05	0
133	Proline (%)	0652	1.3200	0.00000	1.3756	0.06071	0.02685	24	-0.92	133.00	0
133	Proline (%)	0859	1.3245	0.01300	1.3756	0.06071	0.02685	24	-0.84	133.00	0
133	Proline (%)	0939	1.3450	0.01000	1.3756	0.06071	0.02685	24	-0.50	133.05	0
133	Proline (%)	0504	1.3500	0.00000	1.3756	0.06071	0.02685	24	-0.42	133.00	0
133	Proline (%)	0610	1.3500	0.02000	1.3756	0.06071	0.02685	24	-0.42	133.05	0
133	Proline (%)	0644	1.3590	0.01000	1.3756	0.06071	0.02685	24	-0.27	133.00	0
133	Proline (%)	0571	1.3645	0.04100	1.3756	0.06071	0.02685	24	-0.18	133.00	0
133	Proline (%)	0968	1.3670	0.00200	1.3756	0.06071	0.02685	24	-0.14	133.00	0
133	Proline (%)	0227	1.3700	0.04000	1.3756	0.06071	0.02685	24	-0.09	133.00	0
133	Proline (%)	0918	1.3800	0.14350	1.3756	0.06071	0.02685	24	0.07	133.00	0
133	Proline (%)	0941	1.3950	0.01000	1.3756	0.06071	0.02685	24	0.32	133.00	0
133	Proline (%)	0676	1.3950	0.03000	1.3756	0.06071	0.02685	24	0.32	133.05	0
133	Proline (%)	0675	1.4000	0.02000	1.3756	0.06071	0.02685	24	0.40	133.00	0
133	Proline (%)	0354	1.4100	0.05000	1.3756	0.06071	0.02685	24	0.57	133.00	0
133	Proline (%)	0626	1.4160	0.03600	1.3756	0.06071	0.02685	24	0.67	133.05	0
133	Proline (%)	2059	1.4300	0.01000	1.3756	0.06071	0.02685	24	0.90	133.00	0
133	Proline (%)	0910	1.4500	0.00000	1.3756	0.06071	0.02685	24	1.23	133.00	0
133	Proline (%)	0904	1.4500	0.02000	1.3756	0.06071	0.02685	24	1.23	133.99	0
133	Proline (%)	0619	1.4700	0.04000	1.3756	0.06071	0.02685	24	1.55	133.00	0
133	Proline (%)	0226	2.8650	0.07000	1.3756	0.06071	0.02685	24	24.53	133.99	0
134	Serine (%)	0504	0.72500	0.05000	0.86532	0.03800	0.01785	25	-3.69	134.00	0
134	Serine (%)	0676	0.81000	0.04000	0.86532	0.03800	0.01785	25	-1.46	134.05	0
134	Serine (%)	0098	0.83100	0.01200	0.86532	0.03800	0.01785	25	-0.90	134.02	0
134	Serine (%)	0644	0.83500	0.00800	0.86532	0.03800	0.01785	25	-0.80	134.00	0
134	Serine (%)	0610	0.83500	0.01000	0.86532	0.03800	0.01785	25	-0.80	134.05	0
134	Serine (%)	0684	0.84000	0.03800	0.86532	0.03800	0.01785	25	-0.67	134.00	0
134	Serine (%)	0626	0.84450	0.00700	0.86532	0.03800	0.01785	25	-0.55	134.05	0
134	Serine (%)	0652	0.84500	0.01000	0.86532	0.03800	0.01785	25	-0.53	134.00	0
134	Serine (%)	0904	0.85000	0.00000	0.86532	0.03800	0.01785	25	-0.40	134.99	0
134	Serine (%)	0872	0.85050	0.00900	0.86532	0.03800	0.01785	25	-0.39	134.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
134	Serine (%)	2059	0.85200	0.00200	0.86532	0.03800	0.01785	25	-0.35	134.00	0
134	Serine (%)	0619	0.85300	0.02400	0.86532	0.03800	0.01785	25	-0.32	134.00	0
134	Serine (%)	0227	0.85500	0.01000	0.86532	0.03800	0.01785	25	-0.27	134.00	0
134	Serine (%)	0571	0.86500	0.01200	0.86532	0.03800	0.01785	25	-0.01	134.00	0
134	Serine (%)	0941	0.86500	0.01000	0.86532	0.03800	0.01785	25	-0.01	134.00	0
134	Serine (%)	0968	0.87150	0.01300	0.86532	0.03800	0.01785	25	0.16	134.00	0
134	Serine (%)	0859	0.87650	0.00300	0.86532	0.03800	0.01785	25	0.29	134.00	0
134	Serine (%)	0675	0.88000	0.02000	0.86532	0.03800	0.01785	25	0.39	134.00	0
134	Serine (%)	0910	0.88500	0.03000	0.86532	0.03800	0.01785	25	0.52	134.00	0
134	Serine (%)	0171	0.90000	0.00000	0.86532	0.03800	0.01785	25	0.91	134.00	0
134	Serine (%)	0939	0.90000	0.00000	0.86532	0.03800	0.01785	25	0.91	134.05	0
134	Serine (%)	0918	0.91510	0.02420	0.86532	0.03800	0.01785	25	1.31	134.00	0
134	Serine (%)	0354	0.92100	0.01200	0.86532	0.03800	0.01785	25	1.47	134.00	0
134	Serine (%)	0148	0.92400	0.02200	0.86532	0.03800	0.01785	25	1.54	134.05	0
134	Serine (%)	0226	2.1400	0.08000	0.86532	0.03800	0.01785	25	33.54	134.99	0
135	Threonine (%)	0148	0.63650	0.00700	0.68499	0.02057	0.00788	25	-2.36	135.05	0
135	Threonine (%)	0626	0.65800	0.01600	0.68499	0.02057	0.00788	25	-1.31	135.05	0
135	Threonine (%)	0652	0.66000	0.00000	0.68499	0.02057	0.00788	25	-1.21	135.00	0
135	Threonine (%)	0939	0.66500	0.01000	0.68499	0.02057	0.00788	25	-0.97	135.05	0
135	Threonine (%)	0504	0.67000	0.00000	0.68499	0.02057	0.00788	25	-0.73	135.00	0
135	Threonine (%)	0684	0.67050	0.01900	0.68499	0.02057	0.00788	25	-0.70	135.00	0
135	Threonine (%)	0619	0.67150	0.02300	0.68499	0.02057	0.00788	25	-0.66	135.00	0
135	Threonine (%)	0941	0.67500	0.01000	0.68499	0.02057	0.00788	25	-0.49	135.00	0
135	Threonine (%)	0859	0.67600	0.00000	0.68499	0.02057	0.00788	25	-0.44	135.00	0
135	Threonine (%)	2059	0.67750	0.00100	0.68499	0.02057	0.00788	25	-0.36	135.00	0
135	Threonine (%)	0610	0.68000	0.00000	0.68499	0.02057	0.00788	25	-0.24	135.05	0
135	Threonine (%)	0571	0.68200	0.00000	0.68499	0.02057	0.00788	25	-0.15	135.00	0
135	Threonine (%)	0644	0.68300	0.00800	0.68499	0.02057	0.00788	25	-0.10	135.00	0
135	Threonine (%)	0098	0.68800	0.03200	0.68499	0.02057	0.00788	25	0.15	135.02	0
135	Threonine (%)	0227	0.69000	0.00000	0.68499	0.02057	0.00788	25	0.24	135.00	0
135	Threonine (%)	0675	0.69000	0.00000	0.68499	0.02057	0.00788	25	0.24	135.00	0
135	Threonine (%)	0676	0.69000	0.00000	0.68499	0.02057	0.00788	25	0.24	135.05	0
135	Threonine (%)	0872	0.69200	0.00400	0.68499	0.02057	0.00788	25	0.34	135.00	0
135	Threonine (%)	0910	0.69500	0.01000	0.68499	0.02057	0.00788	25	0.49	135.00	0
135	Threonine (%)	0904	0.69500	0.01000	0.68499	0.02057	0.00788	25	0.49	135.99	0
135	Threonine (%)	0968	0.69950	0.00100	0.68499	0.02057	0.00788	25	0.71	135.00	0
135	Threonine (%)	0171	0.71500	0.01000	0.68499	0.02057	0.00788	25	1.46	135.00	0
135	Threonine (%)	0354	0.73450	0.00100	0.68499	0.02057	0.00788	25	2.41	135.00	0
135	Threonine (%)	0918	0.73700	0.02500	0.68499	0.02057	0.00788	25	2.53	135.00	0
135	Threonine (%)	0226	1.4950	0.01000	0.68499	0.02057	0.00788	25	39.38	135.99	0
136	Tryptophan (%)	0859	0.17550	0.00500	0.19947	0.01935	0.00697	15	-1.24	136.03	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
136	Tryptophan (%)	0904	0.18000	0.00000	0.19947	0.01935	0.00697	15	-1.01	136.99	0
136	Tryptophan (%)	0941	0.18500	0.01000	0.19947	0.01935	0.00697	15	-0.75	136.00	0
136	Tryptophan (%)	0872	0.18650	0.00300	0.19947	0.01935	0.00697	15	-0.67	136.03	0
136	Tryptophan (%)	0098	0.18750	0.00300	0.19947	0.01935	0.00697	15	-0.62	136.02	0
136	Tryptophan (%)	2059	0.18900	0.00000	0.19947	0.01935	0.00697	15	-0.54	136.01	0
136	Tryptophan (%)	0571	0.19100	0.00000	0.19947	0.01935	0.00697	15	-0.44	136.03	0
136	Tryptophan (%)	0619	0.19350	0.01700	0.19947	0.01935	0.00697	15	-0.31	136.00	0
136	Tryptophan (%)	0644	0.20000	0.00000	0.19947	0.01935	0.00697	15	0.03	136.01	0
136	Tryptophan (%)	0918	0.20700	0.02360	0.19947	0.01935	0.00697	15	0.39	136.00	0
136	Tryptophan (%)	0610	0.21000	0.00000	0.19947	0.01935	0.00697	15	0.54	136.99	0
136	Tryptophan (%)	0504	0.21000	0.02000	0.19947	0.01935	0.00697	15	0.54	136.99	0
136	Tryptophan (%)	0227	0.22000	0.00000	0.19947	0.01935	0.00697	15	1.06	136.00	0
136	Tryptophan (%)	0684	0.40850	0.01300	0.19947	0.01935	0.00697	15	10.80	136.00	0
136	Tryptophan (%)	0226	0.41500	0.01000	0.19947	0.01935	0.00697	15	11.14	136.99	0
137	Tyrosine (%)	0675	0.56000	0.00000	0.63746	0.05505	0.02767	21	-1.41	137.00	0
137	Tyrosine (%)	0626	0.56700	0.02000	0.63746	0.05505	0.02767	21	-1.28	137.05	0
137	Tyrosine (%)	0227	0.57500	0.01000	0.63746	0.05505	0.02767	21	-1.13	137.00	0
137	Tyrosine (%)	0504	0.57500	0.05000	0.63746	0.05505	0.02767	21	-1.13	137.00	0
137	Tyrosine (%)	0684	0.60250	0.02700	0.63746	0.05505	0.02767	21	-0.64	137.00	0
137	Tyrosine (%)	0910	0.60500	0.01000	0.63746	0.05505	0.02767	21	-0.59	137.00	0
137	Tyrosine (%)	0148	0.60750	0.02500	0.63746	0.05505	0.02767	21	-0.54	137.05	0
137	Tyrosine (%)	0171	0.62000	0.00000	0.63746	0.05505	0.02767	21	-0.32	137.00	0
137	Tyrosine (%)	0918	0.62515	0.00010	0.63746	0.05505	0.02767	21	-0.22	137.00	0
137	Tyrosine (%)	0098	0.63050	0.03500	0.63746	0.05505	0.02767	21	-0.13	137.02	0
137	Tyrosine (%)	0941	0.64500	0.09000	0.63746	0.05505	0.02767	21	0.14	137.00	0
137	Tyrosine (%)	0644	0.64550	0.01100	0.63746	0.05505	0.02767	21	0.15	137.00	0
137	Tyrosine (%)	0354	0.64650	0.00300	0.63746	0.05505	0.02767	21	0.16	137.00	0
137	Tyrosine (%)	0968	0.65250	0.02100	0.63746	0.05505	0.02767	21	0.27	137.00	0
137	Tyrosine (%)	0676	0.65500	0.03000	0.63746	0.05505	0.02767	21	0.32	137.05	0
137	Tyrosine (%)	0872	0.66950	0.00900	0.63746	0.05505	0.02767	21	0.58	137.00	0
137	Tyrosine (%)	0904	0.67000	0.04000	0.63746	0.05505	0.02767	21	0.59	137.99	0
137	Tyrosine (%)	2059	0.67500	0.00000	0.63746	0.05505	0.02767	21	0.68	137.00	0
137	Tyrosine (%)	0610	0.72000	0.02000	0.63746	0.05505	0.02767	21	1.50	137.05	0
137	Tyrosine (%)	0939	0.80500	0.01000	0.63746	0.05505	0.02767	21	3.04	137.05	0
137	Tyrosine (%)	0226	1.5150	0.17000	0.63746	0.05505	0.02767	21	15.94	137.99	0
138	Valine (%)	0148	0.74300	0.01400	0.87817	0.05096	0.02565	24	-2.65	138.05	0
138	Valine (%)	0684	0.77450	0.02100	0.87817	0.05096	0.02565	24	-2.03	138.00	0
138	Valine (%)	0859	0.81200	0.01200	0.87817	0.05096	0.02565	24	-1.30	138.00	0
138	Valine (%)	0939	0.82500	0.01000	0.87817	0.05096	0.02565	24	-1.04	138.05	0
138	Valine (%)	0171	0.83000	0.00000	0.87817	0.05096	0.02565	24	-0.95	138.00	0
138	Valine (%)	0652	0.85000	0.02000	0.87817	0.05096	0.02565	24	-0.55	138.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
138	Valine (%)	0619	0.85200	0.01800	0.87817	0.05096	0.02565	24	-0.51	138.00	0
138	Valine (%)	0968	0.85350	0.00100	0.87817	0.05096	0.02565	24	-0.48	138.00	0
138	Valine (%)	0918	0.86685	0.00770	0.87817	0.05096	0.02565	24	-0.22	138.00	0
138	Valine (%)	0610	0.87000	0.00000	0.87817	0.05096	0.02565	24	-0.16	138.05	0
138	Valine (%)	0626	0.87050	0.01900	0.87817	0.05096	0.02565	24	-0.15	138.05	0
138	Valine (%)	0941	0.87500	0.03000	0.87817	0.05096	0.02565	24	-0.06	138.00	0
138	Valine (%)	0872	0.87650	0.01700	0.87817	0.05096	0.02565	24	-0.03	138.00	0
138	Valine (%)	2059	0.88850	0.00100	0.87817	0.05096	0.02565	24	0.20	138.00	0
138	Valine (%)	0571	0.88900	0.00000	0.87817	0.05096	0.02565	24	0.21	138.00	0
138	Valine (%)	0098	0.89150	0.00100	0.87817	0.05096	0.02565	24	0.26	138.02	0
138	Valine (%)	0354	0.90600	0.00400	0.87817	0.05096	0.02565	24	0.55	138.00	0
138	Valine (%)	0644	0.90700	0.01000	0.87817	0.05096	0.02565	24	0.57	138.00	0
138	Valine (%)	0910	0.91000	0.02000	0.87817	0.05096	0.02565	24	0.62	138.00	0
138	Valine (%)	0227	0.92500	0.01000	0.87817	0.05096	0.02565	24	0.92	138.00	0
138	Valine (%)	0675	0.92500	0.01000	0.87817	0.05096	0.02565	24	0.92	138.00	0
138	Valine (%)	0504	0.94000	0.04000	0.87817	0.05096	0.02565	24	1.21	138.00	0
138	Valine (%)	0676	0.96000	0.02000	0.87817	0.05096	0.02565	24	1.61	138.05	0
138	Valine (%)	0226	1.6950	0.33000	0.87817	0.05096	0.02565	24	16.03	138.99	0
139	Taurine (%)	0941	0.11500	0.01000				2		139.00	0
139	Taurine (%)	0504	0.15000	0.00000				2		139.00	0
139	Taurine (%)	0098	0.00000	0.00000				2		139.02	4
160	Fructose (%)	0619	0.15700	0.00400	0.42800	0.45925	0.01850	4	-1.01	160.99	0
160	Fructose (%)	0861	0.20000	0.00000	0.42800	0.45925	0.01850	4	0.02	160.99	0
160	Fructose (%)	0921	0.24000	0.04000	0.42800	0.45925	0.01850	4	0.99	160.99	0
160	Fructose (%)	0889	1.1150	0.03000	0.42800	0.45925	0.01850	4	22.07	160.99	0
162	Glucose (%)	0889	0.14000	0.00000	0.38213	0.27782	0.06075	4	-0.61	162.99	0
162	Glucose (%)	0619	0.15850	0.00300	0.38213	0.27782	0.06075	4	-0.55	162.99	0
162	Glucose (%)	0921	0.53000	0.24000	0.38213	0.27782	0.06075	4	0.62	162.99	0
162	Glucose (%)	0629	0.70000	0.00000	0.38213	0.27782	0.06075	4	1.15	162.99	0
162	Glucose (%)	0861	0.00000	0.00000	0.38213	0.27782	0.06075	4		162.99	4
163	Lactose (%)	0619	0.00000	0.00000				0		163.99	4
163	Lactose (%)	0861	0.00000	0.00000				0		163.99	4
164	Maltose (%)	0921	0.75000	0.04000				1		164.99	0
164	Maltose (%)	0861	0.00000	0.00000				1		164.99	4
165	Sucrose (%)	0889	0.15000	0.02000	1.1090	0.53814	0.01800	5	-22.17	165.99	0
165	Sucrose (%)	0629	1.2750	0.05000	1.1090	0.53814	0.01800	5	-1.36	165.99	0
165	Sucrose (%)	0227	1.3450	0.01000	1.1090	0.53814	0.01800	5	-0.07	165.99	0
165	Sucrose (%)	0619	1.3750	0.01000	1.1090	0.53814	0.01800	5	0.49	165.99	0
165	Sucrose (%)	0861	1.4000	0.00000	1.1090	0.53814	0.01800	5	0.95	165.99	0
166	Raffinose (%)	0227	0.28000	0.02000				2		166.99	0
166	Raffinose (%)	0619	0.36500	0.00200				2		166.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
167	Stachyose (%)	0227	0.20500	0.01000				2		167.99	0
167	Stachyose (%)	0619	0.29900	0.00000				2		167.99	0
314	Semduramicin sodium (mg/kg (ppm))	0218	0.00000	0.00000				0		314.04	4
317	Maduramicin (mg/kg (ppm))	0218	0.00000	0.00000				0		317.05	4
345	Amprolium (mg/kg (ppm))	0218	1.3000	0.00000	14.643	22.238	1.6500	5	-0.54	345.02	0
345	Amprolium (mg/kg (ppm))	0875	1.5500	0.10000	14.643	22.238	1.6500	5	-0.50	345.04	0
345	Amprolium (mg/kg (ppm))	2053	1.8200	0.06000	14.643	22.238	1.6500	5	-0.46	345.03	0
345	Amprolium (mg/kg (ppm))	0036	15.650	0.50000	14.643	22.238	1.6500	5	1.50	345.00	0
345	Amprolium (mg/kg (ppm))	0038	52.895	7.5900	14.643	22.238	1.6500	5	6.78	345.02	0
351	Chlortetracycline (mg/kg (ppm))	0038	0.49020	0.20360				2		351.03	0
351	Chlortetracycline (mg/kg (ppm))	2127	1.0200	0.20000				2		351.05	0
354	Decoquinat e (mg/kg (ppm))	0964	0.77130	0.11060	0.93809	0.13093	0.05793	9	-1.27	354.02	0
354	Decoquinat e (mg/kg (ppm))	2127	0.79900	0.05200	0.93809	0.13093	0.05793	9	-1.06	354.04	0
354	Decoquinat e (mg/kg (ppm))	2135	0.89600	0.00780	0.93809	0.13093	0.05793	9	-0.32	354.04	0
354	Decoquinat e (mg/kg (ppm))	0610	0.90500	0.01000	0.93809	0.13093	0.05793	9	-0.25	354.04	0
354	Decoquinat e (mg/kg (ppm))	0875	0.91500	0.05000	0.93809	0.13093	0.05793	9	-0.18	354.04	0
354	Decoquinat e (mg/kg (ppm))	0038	0.97200	0.07600	0.93809	0.13093	0.05793	9	0.26	354.02	0
354	Decoquinat e (mg/kg (ppm))	0218	1.0000	0.00000	0.93809	0.13093	0.05793	9	0.47	354.01	0
354	Decoquinat e (mg/kg (ppm))	0036	1.0500	0.10000	0.93809	0.13093	0.05793	9	0.85	354.02	0
354	Decoquinat e (mg/kg (ppm))	0027	1.4285	0.11500	0.93809	0.13093	0.05793	9	3.75	354.01	0
355	Erythromycin (mg/kg (ppm))	0218	0.00000	0.00000				0		355.03	4
357	Ethoxyquin (mg/kg (ppm))	0227	1.0000	0.00000				2		357.01	0
357	Ethoxyquin (mg/kg (ppm))	0939	1.1050	0.05000				2		357.01	0
361	Lasalocid sodium (mg/kg (ppm))	2127	0.18600	0.05800	0.81936	0.10945	0.04004	9	-5.79	361.05	0
361	Lasalocid sodium (mg/kg (ppm))	0610	0.72500	0.07000	0.81936	0.10945	0.04004	9	-0.86	361.05	0
361	Lasalocid sodium (mg/kg (ppm))	0218	0.79000	0.02000	0.81936	0.10945	0.04004	9	-0.27	361.05	0
361	Lasalocid sodium (mg/kg (ppm))	2149	0.80000	0.00000	0.81936	0.10945	0.04004	9	-0.18	361.03	0
361	Lasalocid sodium (mg/kg (ppm))	0875	0.80000	0.00000	0.81936	0.10945	0.04004	9	-0.18	361.05	0
361	Lasalocid sodium (mg/kg (ppm))	2135	0.84560	0.06140	0.81936	0.10945	0.04004	9	0.24	361.05	0
361	Lasalocid sodium (mg/kg (ppm))	0027	0.88075	0.07270	0.81936	0.10945	0.04004	9	0.56	361.02	0
361	Lasalocid sodium (mg/kg (ppm))	0038	0.89415	0.03830	0.81936	0.10945	0.04004	9	0.68	361.03	0
361	Lasalocid sodium (mg/kg (ppm))	2053	1.1200	0.04000	0.81936	0.10945	0.04004	9	2.75	361.04	0
363	Lincomycin (mg/kg (ppm))	0218	0.00000	0.00000				0		363.03	4
365	Monensin (mg/kg (ppm))	0218	0.00000	0.00000				0		365.05	4
370	Novobiocin (mg/kg (ppm))	0218	0.00000	0.00000				0		370.03	4
373	Oxytetracycline (mg/kg (ppm))	0038	1.5805	2.3990				3		373.04	0
373	Oxytetracycline (mg/kg (ppm))	0218	1.6500	0.10000				3		373.03	0
373	Oxytetracycline (mg/kg (ppm))	2053	2.9500	0.10000				3		373.05	0
379	Salinomycin (mg/kg (ppm))	0218	0.00000	0.00000				0		379.05	4
388	Tylosin (mg/kg (ppm))	0218	0.45000	0.08000				3		388.05	0
388	Tylosin (mg/kg (ppm))	0038	0.81300	0.04200				3		388.03	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
388	Tylosin (mg/kg (ppm))	0875	1.2000	0.40000				3		388.05	0
389	Virginiamycin (mg/kg (ppm))	0218	0.00000	0.00000				0		389.03	4
391	Narasin (mg/kg (ppm))	0218	0.00000	0.00000				0		391.03	4
400	Water activity (Units)	0843	0.57850	0.00500	0.59371	0.01816	0.00409	8	-0.84	400.01	0
400	Water activity (Units)	0942	0.57900	0.00200	0.59371	0.01816	0.00409	8	-0.81	400.01	0
400	Water activity (Units)	0407	0.58000	0.00020	0.59371	0.01816	0.00409	8	-0.75	400.01	0
400	Water activity (Units)	0843	0.58600	0.00400	0.59371	0.01816	0.00409	8	-0.42	400.99	0
400	Water activity (Units)	0589	0.58850	0.00300	0.59371	0.01816	0.00409	8	-0.29	400.01	0
400	Water activity (Units)	2109	0.61055	0.00030	0.59371	0.01816	0.00409	8	0.93	400.01	0
400	Water activity (Units)	0552	0.61260	0.00120	0.59371	0.01816	0.00409	8	1.04	400.01	0
400	Water activity (Units)	2073	0.61450	0.01700	0.59371	0.01816	0.00409	8	1.15	400.99	0
516	Arsenic, total (mg / kg (ppm))	0199	0.12000	0.00000	0.14445	0.02069	0.01210	9	-1.18	516.53	0
516	Arsenic, total (mg / kg (ppm))	0096	0.12680	0.00720	0.14445	0.02069	0.01210	9	-0.85	516.52	0
516	Arsenic, total (mg / kg (ppm))	0171	0.13000	0.00000	0.14445	0.02069	0.01210	9	-0.70	516.00	0
516	Arsenic, total (mg / kg (ppm))	0227	0.13300	0.00000	0.14445	0.02069	0.01210	9	-0.55	516.53	0
516	Arsenic, total (mg / kg (ppm))	0553	0.14550	0.01900	0.14445	0.02069	0.01210	9	0.05	516.53	0
516	Arsenic, total (mg / kg (ppm))	0918	0.15425	0.00070	0.14445	0.02069	0.01210	9	0.47	516.53	0
516	Arsenic, total (mg / kg (ppm))	0910	0.15500	0.05000	0.14445	0.02069	0.01210	9	0.51	516.52	0
516	Arsenic, total (mg / kg (ppm))	0021	0.16000	0.02000	0.14445	0.02069	0.01210	9	0.75	516.53	0
516	Arsenic, total (mg / kg (ppm))	0098	0.21800	0.01200	0.14445	0.02069	0.01210	9	3.56	516.53	0
516	Arsenic, total (mg / kg (ppm))	0021	0.00000	0.00000	0.14445	0.02069	0.01210	9		516.43	4
518	Cadmium (mg / kg (ppm))	0619	0.04950	0.00100	0.06159	0.00642	0.00290	11	-1.88	518.41	0
518	Cadmium (mg / kg (ppm))	0021	0.05500	0.01000	0.06159	0.00642	0.00290	11	-1.03	518.53	0
518	Cadmium (mg / kg (ppm))	0227	0.05650	0.00300	0.06159	0.00642	0.00290	11	-0.79	518.53	0
518	Cadmium (mg / kg (ppm))	0910	0.06000	0.00000	0.06159	0.00642	0.00290	11	-0.25	518.52	0
518	Cadmium (mg / kg (ppm))	0199	0.06000	0.00000	0.06159	0.00642	0.00290	11	-0.25	518.53	0
518	Cadmium (mg / kg (ppm))	0096	0.06210	0.00400	0.06159	0.00642	0.00290	11	0.08	518.52	0
518	Cadmium (mg / kg (ppm))	0918	0.06245	0.00270	0.06159	0.00642	0.00290	11	0.13	518.53	0
518	Cadmium (mg / kg (ppm))	0171	0.06500	0.01000	0.06159	0.00642	0.00290	11	0.53	518.41	0
518	Cadmium (mg / kg (ppm))	0553	0.06550	0.00120	0.06159	0.00642	0.00290	11	0.61	518.53	0
518	Cadmium (mg / kg (ppm))	0098	0.06900	0.00000	0.06159	0.00642	0.00290	11	1.15	518.53	0
518	Cadmium (mg / kg (ppm))	0968	0.07000	0.00000	0.06159	0.00642	0.00290	11	1.31	518.43	0
518	Cadmium (mg / kg (ppm))	0021	0.00000	0.00000	0.06159	0.00642	0.00290	11		518.43	4
520	Chromium (mg / kg (ppm))	0918	1.7019	0.36150	7.6079	4.5250	1.8642	12	-1.31	520.53	0
520	Chromium (mg / kg (ppm))	0199	3.4900	0.08000	7.6079	4.5250	1.8642	12	-0.91	520.53	0
520	Chromium (mg / kg (ppm))	0968	3.8300	0.00000	7.6079	4.5250	1.8642	12	-0.83	520.43	0
520	Chromium (mg / kg (ppm))	0619	4.0450	0.17000	7.6079	4.5250	1.8642	12	-0.79	520.41	0
520	Chromium (mg / kg (ppm))	0171	5.8000	0.40000	7.6079	4.5250	1.8642	12	-0.40	520.41	0
520	Chromium (mg / kg (ppm))	0510	7.2950	0.35000	7.6079	4.5250	1.8642	12	-0.07	520.43	0
520	Chromium (mg / kg (ppm))	2053	7.6150	0.65000	7.6079	4.5250	1.8642	12	0.00	520.42	0
520	Chromium (mg / kg (ppm))	0553	9.5250	1.7500	7.6079	4.5250	1.8642	12	0.42	520.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
520	Chromium (mg / kg (ppm))	0693	9.6400	4.1200	7.6079	4.5250	1.8642	12	0.45	520.42	0
520	Chromium (mg / kg (ppm))	0096	11.311	8.7819	7.6079	4.5250	1.8642	12	0.82	520.42	0
520	Chromium (mg / kg (ppm))	0096	12.646	1.7065	7.6079	4.5250	1.8642	12	1.11	520.52	0
520	Chromium (mg / kg (ppm))	0021	15.000	4.0000	7.6079	4.5250	1.8642	12	1.63	520.53	0
520	Chromium (mg / kg (ppm))	0021	0.00000	0.00000	7.6079	4.5250	1.8642	12		520.43	4
526	Lead (mg / kg (ppm))	0199	0.17000	0.00000	0.19406	0.01551	0.00877	9	-1.55	526.53	0
526	Lead (mg / kg (ppm))	0227	0.18000	0.00200	0.19406	0.01551	0.00877	9	-0.91	526.53	0
526	Lead (mg / kg (ppm))	0096	0.18095	0.00750	0.19406	0.01551	0.00877	9	-0.85	526.52	0
526	Lead (mg / kg (ppm))	0021	0.19500	0.01000	0.19406	0.01551	0.00877	9	0.06	526.53	0
526	Lead (mg / kg (ppm))	0171	0.20000	0.00000	0.19406	0.01551	0.00877	9	0.38	526.41	0
526	Lead (mg / kg (ppm))	0910	0.20000	0.02000	0.19406	0.01551	0.00877	9	0.38	526.52	0
526	Lead (mg / kg (ppm))	0918	0.20130	0.02240	0.19406	0.01551	0.00877	9	0.47	526.53	0
526	Lead (mg / kg (ppm))	0098	0.20900	0.00400	0.19406	0.01551	0.00877	9	0.96	526.53	0
526	Lead (mg / kg (ppm))	0553	0.20950	0.01300	0.19406	0.01551	0.00877	9	1.00	526.53	0
526	Lead (mg / kg (ppm))	0619	0.00000	0.00000	0.19406	0.01551	0.00877	9		526.41	4
526	Lead (mg / kg (ppm))	0021	0.00000	0.00000	0.19406	0.01551	0.00877	9		526.43	4
529	Mercury (µg / kg (ppb))	0941	1.2000	0.00000				1		529.99	0
539	Nickel (mg / kg (ppm))	0941	1.4150	0.01000	2.4526	0.90611	0.18661	7	-1.15	539.53	0
539	Nickel (mg / kg (ppm))	0918	1.4657	0.16530	2.4526	0.90611	0.18661	7	-1.09	539.53	0
539	Nickel (mg / kg (ppm))	0171	2.3500	0.10000	2.4526	0.90611	0.18661	7	-0.11	539.41	0
539	Nickel (mg / kg (ppm))	2053	2.4200	0.14000	2.4526	0.90611	0.18661	7	-0.04	539.42	0
539	Nickel (mg / kg (ppm))	0096	2.9774	0.23100	2.4526	0.90611	0.18661	7	0.58	539.52	0
539	Nickel (mg / kg (ppm))	0553	2.9900	0.56000	2.4526	0.90611	0.18661	7	0.59	539.53	0
539	Nickel (mg / kg (ppm))	0021	3.5500	0.10000	2.4526	0.90611	0.18661	7	1.21	539.53	0
539	Nickel (mg / kg (ppm))	0021	1.6000	3.2000	2.4526	0.90611	0.18661	7		539.43	4
702	Butyric Acid (4:0) (%)	0619	0.01095	0.00010				1		702.00	0
704	Caproic Acid (6:0) (%)	0619	0.00085	0.00010				1		704.00	0
706	Caprylic acid (8:0) (%) (w/w)	0619	0.01955	0.00070				1		706.01	0
708	Capric acid (10:0) (%) (w/w)	0619	0.02390	0.00060				1		708.01	0
710	Lauric Acid (12:0) (%) (w/w)	0619	0.04810	0.00140				2		710.01	0
710	Lauric Acid (12:0) (%) (w/w)	0676	0.05350	0.00100				2		710.99	0
714	Myristic Acid (14:0) (%) (w/w)	0619	0.03335	0.00070				1		714.01	0
716	Palmitic Acid (16:0) (%) (w/w)	0619	0.77940	0.00220				1		716.01	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0676	0.01100	0.00000				2		718.99	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0619	0.02730	0.00040				2		718.01	0
722	Stearic Acid (18:0) (%) (w/w)	0619	0.08110	0.00120				1		722.01	0
724	Oleic Acid (9c-18:1) (%) (w/w)	0619	1.2525	0.00700				1		724.01	0
726	Linoleic Acid (9c,12c-18:2) (%) (w/w)	0619	1.8142	0.02070				2		726.01	0
726	Linoleic Acid (9c,12c-18:2) (%) (w/w)	0676	2.4275	0.01500				2		726.99	0
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (‘	0619	0.77615	0.00130				2		728.01	0
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (‘	0676	0.80700	0.01200				2		728.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
730	Arachidic Acid (20:0) (% (w/w))	0619	0.03715	0.00070				1	730.01	0	
732	Gondoic Acid (11c-20:1) (% (w/w))	0619	0.03615	0.00090				1	732.01	0	
736	Arachidonic Acid (5c,8c,11c,14c-20:4) (%)	0619	0.00000	0.00000				0	736.01	4	
736	Arachidonic Acid (5c,8c,11c,14c-20:4) (%)	0676	0.00000	0.00000				0	736.99	4	
738	Mead Acid (11c,14c,17c-20:3) (% (w/w))	0619	0.00000	0.00000				0	738.01	4	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c-20:5) (%)	0676	0.04200	0.00000				1	740.99	0	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c-20:5) (%)	0619	0.00000	0.00000				1	740.01	4	
744	Erucic Acid (13c-22:1) (% (w/w))	0619	0.00000	0.00000				0	744.01	4	
746	Docosapentaenoic Acid n-3 DPA (DHA) (%)	0619	0.00000	0.00000				0	746.01	4	
746	Docosapentaenoic Acid n-3 DPA (DHA) (%)	0676	0.00000	0.00000				0	746.99	4	
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22:6) (%)	0619	0.00000	0.00000				0	750.01	4	
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22:6) (%)	0676	0.00000	0.00000				0	750.99	4	
752	Nervonic Acid (24:1) isomers (% (w/w))	0619	0.00000	0.00000				0	752.01	4	
754	Total n-3 Polyunsaturated (Omega-3) F&E (%)	0610	0.72500	0.07000	0.81200	0.07791	0.02080	5	-1.55	754.02	0
754	Total n-3 Polyunsaturated (Omega-3) F&E (%)	0629	0.73000	0.00400	0.81200	0.07791	0.02080	5	-1.48	754.99	0
754	Total n-3 Polyunsaturated (Omega-3) F&E (%)	0676	0.85500	0.01000	0.81200	0.07791	0.02080	5	0.30	754.99	0
754	Total n-3 Polyunsaturated (Omega-3) F&E (%)	0675	0.86500	0.01000	0.81200	0.07791	0.02080	5	0.44	754.02	0
754	Total n-3 Polyunsaturated (Omega-3) F&E (%)	0904	0.88500	0.01000	0.81200	0.07791	0.02080	5	0.73	754.01	0
756	Total n-6 Polyunsaturated (Omega-6) F&E (%)	0676	2.4300	0.02000				3		756.99	0
756	Total n-6 Polyunsaturated (Omega-6) F&E (%)	0610	2.5250	0.31000				3		756.02	0
756	Total n-6 Polyunsaturated (Omega-6) F&E (%)	0629	2.5345	0.00100				3		756.99	0
758	Total Saturated Fatty Acids (% (w/w))	0921	17.070	0.00000				1		758.02	0
762	Total Monounsaturated Fatty Acids (% (w/w))	0921	25.460	0.00000				1		762.01	0
766	Total Polyunsaturated Fatty Acids (% (w/w))	0921	57.430	0.00000				1		766.01	0

Notes: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2 and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = Rejected for duplicates too far apart, 2 = Rejected as outlier, 8 = Analyst data exempt and 4 = zeros submitted as values. Robust statistics not used if < 6 labs reporting, in this case the Z Scores may be included for information only (Grey, No Action!). Flag 3 indicates not used in statistics.