



AAFCO
Proficiency Testing Program



Beef Feed

Animal Feed Scheme

Labs Reporting: 193

Test Material Code # 201828

Analyte Proficiency Testing Report

Issue Date : 09/30/2018

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
000	Urea (%)	0278	0.55000	0.10000				1		000.02	0
001	Loss on Drying (%)	0618	6.9750	0.07000	9.0112	0.41420	0.12423	77	-4.92	001.99	0
001	Loss on Drying (%)	2081	7.5750	0.01000	9.0112	0.41420	0.12423	77	-3.47	001.99	0
001	Loss on Drying (%)	0038	7.5800	0.46000	9.0112	0.41420	0.12423	77	-3.46	001.07	0
001	Loss on Drying (%)	0345	7.7400	0.08000	9.0112	0.41420	0.12423	77	-3.07	001.07	0
001	Loss on Drying (%)	0353	8.2550	0.07000	9.0112	0.41420	0.12423	77	-1.83	001.07	0
001	Loss on Drying (%)	0596	8.2750	0.21000	9.0112	0.41420	0.12423	77	-1.78	001.00	0
001	Loss on Drying (%)	2195	8.3550	0.03000	9.0112	0.41420	0.12423	77	-1.58	001.07	0
001	Loss on Drying (%)	0226	8.3900	0.12000	9.0112	0.41420	0.12423	77	-1.50	001.07	0
001	Loss on Drying (%)	2155	8.3900	0.88000	9.0112	0.41420	0.12423	77	-1.50	001.99	0
001	Loss on Drying (%)	0612	8.4900	0.42000	9.0112	0.41420	0.12423	77	-1.26	001.99	0
001	Loss on Drying (%)	0889	8.5100	0.10000	9.0112	0.41420	0.12423	77	-1.21	001.99	0
001	Loss on Drying (%)	2059	8.5300	0.30000	9.0112	0.41420	0.12423	77	-1.16	001.07	0
001	Loss on Drying (%)	0019	8.5500	0.30000	9.0112	0.41420	0.12423	77	-1.11	001.07	0
001	Loss on Drying (%)	0504	8.5850	0.09000	9.0112	0.41420	0.12423	77	-1.03	001.00	0
001	Loss on Drying (%)	0918	8.5850	0.09000	9.0112	0.41420	0.12423	77	-1.03	001.99	0
001	Loss on Drying (%)	0034	8.6000	0.04000	9.0112	0.41420	0.12423	77	-0.99	001.99	0
001	Loss on Drying (%)	0510	8.6000	0.00000	9.0112	0.41420	0.12423	77	-0.99	001.99	0
001	Loss on Drying (%)	0187	8.6800	0.32000	9.0112	0.41420	0.12423	77	-0.80	001.07	0
001	Loss on Drying (%)	0630	8.6850	0.07000	9.0112	0.41420	0.12423	77	-0.79	001.99	0
001	Loss on Drying (%)	0148	8.7600	0.08000	9.0112	0.41420	0.12423	77	-0.61	001.07	0
001	Loss on Drying (%)	0407	8.7644	0.01810	9.0112	0.41420	0.12423	77	-0.60	001.07	0
001	Loss on Drying (%)	0610	8.7800	0.04000	9.0112	0.41420	0.12423	77	-0.56	001.05	0
001	Loss on Drying (%)	0413	8.8500	0.10000	9.0112	0.41420	0.12423	77	-0.39	001.07	0
001	Loss on Drying (%)	0027	8.8900	0.46000	9.0112	0.41420	0.12423	77	-0.29	001.00	0
001	Loss on Drying (%)	2146	8.8900	0.08000	9.0112	0.41420	0.12423	77	-0.29	001.07	0
001	Loss on Drying (%)	0662	8.8993	0.17920	9.0112	0.41420	0.12423	77	-0.27	001.99	0
001	Loss on Drying (%)	0571	8.9050	0.01000	9.0112	0.41420	0.12423	77	-0.26	001.07	0
001	Loss on Drying (%)	0644	8.9200	0.04000	9.0112	0.41420	0.12423	77	-0.22	001.05	0
001	Loss on Drying (%)	0003	8.9400	0.08000	9.0112	0.41420	0.12423	77	-0.17	001.07	0
001	Loss on Drying (%)	0171	8.9400	0.12000	9.0112	0.41420	0.12423	77	-0.17	001.07	0
001	Loss on Drying (%)	0590	8.9800	0.00000	9.0112	0.41420	0.12423	77	-0.08	001.07	0
001	Loss on Drying (%)	0083	8.9850	0.05000	9.0112	0.41420	0.12423	77	-0.06	001.07	0
001	Loss on Drying (%)	0505	9.0000	0.38000	9.0112	0.41420	0.12423	77	-0.03	001.07	0
001	Loss on Drying (%)	0689	9.0000	0.00000	9.0112	0.41420	0.12423	77	-0.03	001.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
001	Loss on Drying (%)	0948	9.0100	0.02000	9.0112	0.41420	0.12423	77	0.00	001.99	0
001	Loss on Drying (%)	0309	9.0275	0.01100	9.0112	0.41420	0.12423	77	0.04	001.00	0
001	Loss on Drying (%)	0035	9.0400	0.08000	9.0112	0.41420	0.12423	77	0.07	001.07	0
001	Loss on Drying (%)	0682	9.0400	0.00000	9.0112	0.41420	0.12423	77	0.07	001.07	0
001	Loss on Drying (%)	0910	9.0500	0.10000	9.0112	0.41420	0.12423	77	0.09	001.99	0
001	Loss on Drying (%)	0619	9.0600	0.08000	9.0112	0.41420	0.12423	77	0.12	001.03	0
001	Loss on Drying (%)	0357	9.0674	0.00240	9.0112	0.41420	0.12423	77	0.14	001.99	0
001	Loss on Drying (%)	0089	9.0800	0.02000	9.0112	0.41420	0.12423	77	0.17	001.07	0
001	Loss on Drying (%)	0878	9.0850	0.13000	9.0112	0.41420	0.12423	77	0.18	001.03	0
001	Loss on Drying (%)	2009	9.0941	0.00280	9.0112	0.41420	0.12423	77	0.20	001.07	0
001	Loss on Drying (%)	0142	9.1100	0.18000	9.0112	0.41420	0.12423	77	0.24	001.07	0
001	Loss on Drying (%)	0876	9.1250	0.15000	9.0112	0.41420	0.12423	77	0.27	001.07	0
001	Loss on Drying (%)	0849	9.1350	0.01000	9.0112	0.41420	0.12423	77	0.30	001.07	0
001	Loss on Drying (%)	0638	9.1700	0.02000	9.0112	0.41420	0.12423	77	0.38	001.07	0
001	Loss on Drying (%)	2012	9.2000	0.02000	9.0112	0.41420	0.12423	77	0.46	001.07	0
001	Loss on Drying (%)	0675	9.2050	0.11000	9.0112	0.41420	0.12423	77	0.47	001.07	0
001	Loss on Drying (%)	2145	9.2150	0.01000	9.0112	0.41420	0.12423	77	0.49	001.07	0
001	Loss on Drying (%)	0139	9.2500	0.02000	9.0112	0.41420	0.12423	77	0.58	001.07	0
001	Loss on Drying (%)	0049	9.2650	0.01000	9.0112	0.41420	0.12423	77	0.61	001.07	0
001	Loss on Drying (%)	0676	9.2700	0.02000	9.0112	0.41420	0.12423	77	0.62	001.99	0
001	Loss on Drying (%)	0693	9.2810	0.61600	9.0112	0.41420	0.12423	77	0.65	001.07	0
001	Loss on Drying (%)	2190	9.2850	0.11000	9.0112	0.41420	0.12423	77	0.66	001.07	0
001	Loss on Drying (%)	0037	9.2850	0.03000	9.0112	0.41420	0.12423	77	0.66	001.99	0
001	Loss on Drying (%)	0186	9.2900	0.20000	9.0112	0.41420	0.12423	77	0.67	001.03	0
001	Loss on Drying (%)	0868	9.2900	0.02000	9.0112	0.41420	0.12423	77	0.67	001.03	0
001	Loss on Drying (%)	0872	9.2900	0.04000	9.0112	0.41420	0.12423	77	0.67	001.07	0
001	Loss on Drying (%)	0843	9.3050	0.31000	9.0112	0.41420	0.12423	77	0.71	001.07	0
001	Loss on Drying (%)	0098	9.3100	0.10000	9.0112	0.41420	0.12423	77	0.72	001.07	0
001	Loss on Drying (%)	2192	9.3100	0.40000	9.0112	0.41420	0.12423	77	0.72	001.07	0
001	Loss on Drying (%)	0581	9.3350	0.07000	9.0112	0.41420	0.12423	77	0.78	001.07	0
001	Loss on Drying (%)	0895	9.3750	0.01000	9.0112	0.41420	0.12423	77	0.88	001.03	0
001	Loss on Drying (%)	0278	9.3750	0.23000	9.0112	0.41420	0.12423	77	0.88	001.07	0
001	Loss on Drying (%)	2126	9.3750	0.01000	9.0112	0.41420	0.12423	77	0.88	001.99	0
001	Loss on Drying (%)	0169	9.4250	0.01000	9.0112	0.41420	0.12423	77	1.00	001.00	0
001	Loss on Drying (%)	0066	9.4490	0.09600	9.0112	0.41420	0.12423	77	1.06	001.07	0
001	Loss on Drying (%)	2109	9.4900	0.02000	9.0112	0.41420	0.12423	77	1.16	001.07	0
001	Loss on Drying (%)	0019	9.4950	0.25000	9.0112	0.41420	0.12423	77	1.17	001.99	0
001	Loss on Drying (%)	0028	9.5050	0.33000	9.0112	0.41420	0.12423	77	1.19	001.07	0
001	Loss on Drying (%)	2053	9.5350	0.01000	9.0112	0.41420	0.12423	77	1.26	001.07	0
001	Loss on Drying (%)	0695	9.6750	0.01000	9.0112	0.41420	0.12423	77	1.60	001.07	0
001	Loss on Drying (%)	0852	9.6750	0.05000	9.0112	0.41420	0.12423	77	1.60	001.99	0
001	Loss on Drying (%)	0218	9.8000	0.20000	9.0112	0.41420	0.12423	77	1.90	001.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
001	Loss on Drying (%)	0074	9.8750	0.15000	9.0112	0.41420	0.12423	77	2.09	001.07	0
001	Loss on Drying (%)	0560	8.7150	1.0300	9.0112	0.41420	0.12423	77	-0.72	001.99	1
001	Loss on Drying (%)	0045	10.340	1.5400	9.0112	0.41420	0.12423	77	3.21	001.07	1
002	Protein, Crude (%)	0874	10.635	0.39000	12.046	0.29606	0.11323	185	-4.77	002.04	0
002	Protein, Crude (%)	0144	10.790	0.02000	12.046	0.29606	0.11323	185	-4.24	002.06	0
002	Protein, Crude (%)	0622	11.062	0.00660	12.046	0.29606	0.11323	185	-3.32	002.05	0
002	Protein, Crude (%)	0853	11.100	0.40000	12.046	0.29606	0.11323	185	-3.20	002.06	0
002	Protein, Crude (%)	0187	11.285	0.01000	12.046	0.29606	0.11323	185	-2.57	002.04	0
002	Protein, Crude (%)	2097	11.393	0.00650	12.046	0.29606	0.11323	185	-2.21	002.05	0
002	Protein, Crude (%)	0505	11.410	0.00000	12.046	0.29606	0.11323	185	-2.15	002.06	0
002	Protein, Crude (%)	2155	11.432	0.35500	12.046	0.29606	0.11323	185	-2.08	002.06	0
002	Protein, Crude (%)	0581	11.475	0.05000	12.046	0.29606	0.11323	185	-1.93	002.06	0
002	Protein, Crude (%)	0968	11.491	0.15900	12.046	0.29606	0.11323	185	-1.88	002.01	0
002	Protein, Crude (%)	2128	11.492	0.23900	12.046	0.29606	0.11323	185	-1.87	002.03	0
002	Protein, Crude (%)	0265	11.500	0.40000	12.046	0.29606	0.11323	185	-1.84	002.06	0
002	Protein, Crude (%)	0676	11.505	0.25000	12.046	0.29606	0.11323	185	-1.83	002.06	0
002	Protein, Crude (%)	0042	11.520	0.22000	12.046	0.29606	0.11323	185	-1.78	002.02	0
002	Protein, Crude (%)	0405	11.540	0.00000	12.046	0.29606	0.11323	185	-1.71	002.04	0
002	Protein, Crude (%)	0345	11.550	0.30000	12.046	0.29606	0.11323	185	-1.68	002.06	0
002	Protein, Crude (%)	0685	11.555	0.01000	12.046	0.29606	0.11323	185	-1.66	002.01	0
002	Protein, Crude (%)	0651	11.572	0.14500	12.046	0.29606	0.11323	185	-1.60	002.05	0
002	Protein, Crude (%)	0539	11.585	0.19000	12.046	0.29606	0.11323	185	-1.56	002.06	0
002	Protein, Crude (%)	0675	11.625	0.05000	12.046	0.29606	0.11323	185	-1.42	002.05	0
002	Protein, Crude (%)	0100	11.640	0.06000	12.046	0.29606	0.11323	185	-1.37	002.06	0
002	Protein, Crude (%)	2144	11.665	0.07000	12.046	0.29606	0.11323	185	-1.29	002.99	0
002	Protein, Crude (%)	0921	11.700	0.00000	12.046	0.29606	0.11323	185	-1.17	002.05	0
002	Protein, Crude (%)	2199	11.700	0.00000	12.046	0.29606	0.11323	185	-1.17	002.05	0
002	Protein, Crude (%)	0956	11.700	0.20000	12.046	0.29606	0.11323	185	-1.17	002.06	0
002	Protein, Crude (%)	2192	11.705	0.15000	12.046	0.29606	0.11323	185	-1.15	002.05	0
002	Protein, Crude (%)	0541	11.735	0.49000	12.046	0.29606	0.11323	185	-1.05	002.06	0
002	Protein, Crude (%)	0910	11.750	0.10000	12.046	0.29606	0.11323	185	-1.00	002.01	0
002	Protein, Crude (%)	2195	11.760	0.18000	12.046	0.29606	0.11323	185	-0.97	002.01	0
002	Protein, Crude (%)	0653	11.765	0.01000	12.046	0.29606	0.11323	185	-0.95	002.01	0
002	Protein, Crude (%)	0036	11.772	0.00070	12.046	0.29606	0.11323	185	-0.93	002.02	0
002	Protein, Crude (%)	0098	11.795	0.03000	12.046	0.29606	0.11323	185	-0.85	002.08	0
002	Protein, Crude (%)	0619	11.800	0.00000	12.046	0.29606	0.11323	185	-0.83	002.05	0
002	Protein, Crude (%)	0047	11.800	0.40000	12.046	0.29606	0.11323	185	-0.83	002.06	0
002	Protein, Crude (%)	2145	11.810	0.08000	12.046	0.29606	0.11323	185	-0.80	002.05	0
002	Protein, Crude (%)	0074	11.810	0.00000	12.046	0.29606	0.11323	185	-0.80	002.06	0
002	Protein, Crude (%)	0881	11.818	0.00900	12.046	0.29606	0.11323	185	-0.77	002.05	0
002	Protein, Crude (%)	0169	11.820	0.06000	12.046	0.29606	0.11323	185	-0.76	002.02	0
002	Protein, Crude (%)	0042	11.820	0.20000	12.046	0.29606	0.11323	185	-0.76	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein, Crude (%)	0914	11.825	0.27000	12.046	0.29606	0.11323	185	-0.75	002.00	0
002	Protein, Crude (%)	0590	11.825	0.21000	12.046	0.29606	0.11323	185	-0.75	002.06	0
002	Protein, Crude (%)	0536	11.830	0.10000	12.046	0.29606	0.11323	185	-0.73	002.03	0
002	Protein, Crude (%)	0049	11.840	0.36000	12.046	0.29606	0.11323	185	-0.70	002.06	0
002	Protein, Crude (%)	0164	11.845	0.19000	12.046	0.29606	0.11323	185	-0.68	002.06	0
002	Protein, Crude (%)	0123	11.875	0.01000	12.046	0.29606	0.11323	185	-0.58	002.06	0
002	Protein, Crude (%)	0868	11.885	0.01000	12.046	0.29606	0.11323	185	-0.54	002.05	0
002	Protein, Crude (%)	0354	11.895	0.05000	12.046	0.29606	0.11323	185	-0.51	002.05	0
002	Protein, Crude (%)	2109	11.895	0.03000	12.046	0.29606	0.11323	185	-0.51	002.05	0
002	Protein, Crude (%)	0294	11.900	0.00000	12.046	0.29606	0.11323	185	-0.49	002.06	0
002	Protein, Crude (%)	0662	11.906	0.18750	12.046	0.29606	0.11323	185	-0.47	002.01	0
002	Protein, Crude (%)	0952	11.910	0.08000	12.046	0.29606	0.11323	185	-0.46	002.05	0
002	Protein, Crude (%)	0032	11.913	0.06900	12.046	0.29606	0.11323	185	-0.45	002.06	0
002	Protein, Crude (%)	0964	11.913	0.07100	12.046	0.29606	0.11323	185	-0.45	002.06	0
002	Protein, Crude (%)	0860	11.913	0.00000	12.046	0.29606	0.11323	185	-0.45	002.01	0
002	Protein, Crude (%)	2009	11.914	0.00280	12.046	0.29606	0.11323	185	-0.45	002.05	0
002	Protein, Crude (%)	0870	11.921	0.14500	12.046	0.29606	0.11323	185	-0.42	002.01	0
002	Protein, Crude (%)	0968	11.922	0.00700	12.046	0.29606	0.11323	185	-0.42	002.06	0
002	Protein, Crude (%)	0939	11.930	0.16000	12.046	0.29606	0.11323	185	-0.39	002.05	0
002	Protein, Crude (%)	0868	11.930	0.06000	12.046	0.29606	0.11323	185	-0.39	002.06	0
002	Protein, Crude (%)	0027	11.933	0.03500	12.046	0.29606	0.11323	185	-0.38	002.06	0
002	Protein, Crude (%)	0139	11.935	0.03000	12.046	0.29606	0.11323	185	-0.37	002.06	0
002	Protein, Crude (%)	2081	11.935	0.17000	12.046	0.29606	0.11323	185	-0.37	002.06	0
002	Protein, Crude (%)	0870	11.940	0.06000	12.046	0.29606	0.11323	185	-0.36	002.06	0
002	Protein, Crude (%)	0872	11.942	0.04300	12.046	0.29606	0.11323	185	-0.35	002.06	0
002	Protein, Crude (%)	0689	11.950	0.10000	12.046	0.29606	0.11323	185	-0.32	002.05	0
002	Protein, Crude (%)	2190	11.950	0.24000	12.046	0.29606	0.11323	185	-0.32	002.05	0
002	Protein, Crude (%)	0175	11.950	0.10000	12.046	0.29606	0.11323	185	-0.32	002.06	0
002	Protein, Crude (%)	0278	11.950	0.10000	12.046	0.29606	0.11323	185	-0.32	002.06	0
002	Protein, Crude (%)	0619	11.950	0.10000	12.046	0.29606	0.11323	185	-0.32	002.06	0
002	Protein, Crude (%)	0878	11.955	0.03000	12.046	0.29606	0.11323	185	-0.31	002.05	0
002	Protein, Crude (%)	2129	11.958	0.09000	12.046	0.29606	0.11323	185	-0.30	002.06	0
002	Protein, Crude (%)	0675	11.960	0.02000	12.046	0.29606	0.11323	185	-0.29	002.06	0
002	Protein, Crude (%)	2188	11.970	0.04000	12.046	0.29606	0.11323	185	-0.26	002.01	0
002	Protein, Crude (%)	0941	11.970	0.08000	12.046	0.29606	0.11323	185	-0.26	002.05	0
002	Protein, Crude (%)	0598	11.980	0.16000	12.046	0.29606	0.11323	185	-0.22	002.06	0
002	Protein, Crude (%)	0066	11.981	0.07900	12.046	0.29606	0.11323	185	-0.22	002.02	0
002	Protein, Crude (%)	0948	11.985	0.01000	12.046	0.29606	0.11323	185	-0.21	002.05	0
002	Protein, Crude (%)	0650	11.990	0.42000	12.046	0.29606	0.11323	185	-0.19	002.06	0
002	Protein, Crude (%)	0202	11.990	0.12000	12.046	0.29606	0.11323	185	-0.19	002.06	0
002	Protein, Crude (%)	2073	11.995	0.13000	12.046	0.29606	0.11323	185	-0.17	002.05	0
002	Protein, Crude (%)	0897	12.000	0.08000	12.046	0.29606	0.11323	185	-0.16	002.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein, Crude (%)	0017	12.000	0.20000	12.046	0.29606	0.11323	185	-0.16	002.06	0
002	Protein, Crude (%)	2012	12.000	0.02000	12.046	0.29606	0.11323	185	-0.16	002.06	0
002	Protein, Crude (%)	0849	12.005	0.01000	12.046	0.29606	0.11323	185	-0.14	002.05	0
002	Protein, Crude (%)	0687	12.010	0.32000	12.046	0.29606	0.11323	185	-0.12	002.06	0
002	Protein, Crude (%)	0857	12.010	0.04000	12.046	0.29606	0.11323	185	-0.12	002.06	0
002	Protein, Crude (%)	2196	12.020	0.00000	12.046	0.29606	0.11323	185	-0.09	002.01	0
002	Protein, Crude (%)	0626	12.036	0.01100	12.046	0.29606	0.11323	185	-0.04	002.06	0
002	Protein, Crude (%)	0859	12.038	0.06900	12.046	0.29606	0.11323	185	-0.03	002.06	0
002	Protein, Crude (%)	2053	12.040	0.02000	12.046	0.29606	0.11323	185	-0.02	002.06	0
002	Protein, Crude (%)	0695	12.041	0.08500	12.046	0.29606	0.11323	185	-0.02	002.06	0
002	Protein, Crude (%)	0610	12.050	0.30000	12.046	0.29606	0.11323	185	0.01	002.01	0
002	Protein, Crude (%)	0098	12.050	0.30000	12.046	0.29606	0.11323	185	0.01	002.06	0
002	Protein, Crude (%)	0520	12.050	0.24000	12.046	0.29606	0.11323	185	0.01	002.06	0
002	Protein, Crude (%)	0038	12.052	0.50800	12.046	0.29606	0.11323	185	0.02	002.06	0
002	Protein, Crude (%)	2022	12.055	0.01000	12.046	0.29606	0.11323	185	0.03	002.05	0
002	Protein, Crude (%)	0504	12.055	0.27000	12.046	0.29606	0.11323	185	0.03	002.06	0
002	Protein, Crude (%)	2150	12.055	0.05000	12.046	0.29606	0.11323	185	0.03	002.06	0
002	Protein, Crude (%)	0171	12.065	0.17000	12.046	0.29606	0.11323	185	0.06	002.06	0
002	Protein, Crude (%)	0589	12.065	0.13000	12.046	0.29606	0.11323	185	0.06	002.06	0
002	Protein, Crude (%)	0918	12.070	0.18000	12.046	0.29606	0.11323	185	0.08	002.06	0
002	Protein, Crude (%)	2181	12.075	0.03000	12.046	0.29606	0.11323	185	0.10	002.05	0
002	Protein, Crude (%)	0009	12.075	0.01000	12.046	0.29606	0.11323	185	0.10	002.06	0
002	Protein, Crude (%)	2006	12.080	0.00000	12.046	0.29606	0.11323	185	0.12	002.05	0
002	Protein, Crude (%)	0693	12.088	0.03200	12.046	0.29606	0.11323	185	0.14	002.06	0
002	Protein, Crude (%)	0652	12.100	0.00000	12.046	0.29606	0.11323	185	0.18	002.01	0
002	Protein, Crude (%)	0108	12.100	0.04000	12.046	0.29606	0.11323	185	0.18	002.06	0
002	Protein, Crude (%)	0861	12.100	0.02000	12.046	0.29606	0.11323	185	0.18	002.06	0
002	Protein, Crude (%)	0895	12.105	0.01000	12.046	0.29606	0.11323	185	0.20	002.05	0
002	Protein, Crude (%)	2012	12.105	0.05000	12.046	0.29606	0.11323	185	0.20	002.05	0
002	Protein, Crude (%)	0939	12.115	0.05000	12.046	0.29606	0.11323	185	0.23	002.06	0
002	Protein, Crude (%)	0003	12.120	0.00000	12.046	0.29606	0.11323	185	0.25	002.06	0
002	Protein, Crude (%)	0357	12.125	0.00000	12.046	0.29606	0.11323	185	0.27	002.06	0
002	Protein, Crude (%)	0880	12.125	0.18800	12.046	0.29606	0.11323	185	0.27	002.06	0
002	Protein, Crude (%)	0354	12.130	0.04000	12.046	0.29606	0.11323	185	0.28	002.06	0
002	Protein, Crude (%)	0504	12.140	0.02000	12.046	0.29606	0.11323	185	0.32	002.04	0
002	Protein, Crude (%)	0142	12.145	0.01000	12.046	0.29606	0.11323	185	0.33	002.06	0
002	Protein, Crude (%)	0571	12.147	0.02500	12.046	0.29606	0.11323	185	0.34	002.06	0
002	Protein, Crude (%)	2059	12.148	0.09060	12.046	0.29606	0.11323	185	0.35	002.06	0
002	Protein, Crude (%)	0045	12.150	0.10000	12.046	0.29606	0.11323	185	0.35	002.06	0
002	Protein, Crude (%)	0226	12.150	0.10000	12.046	0.29606	0.11323	185	0.35	002.06	0
002	Protein, Crude (%)	0910	12.150	0.10000	12.046	0.29606	0.11323	185	0.35	002.06	0
002	Protein, Crude (%)	2168	12.150	0.08000	12.046	0.29606	0.11323	185	0.35	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein, Crude (%)	0674	12.150	0.00000	12.046	0.29606	0.11323	185	0.35	002.05	0
002	Protein, Crude (%)	0208	12.150	0.30000	12.046	0.29606	0.11323	185	0.35	002.06	0
002	Protein, Crude (%)	2089	12.150	0.02000	12.046	0.29606	0.11323	185	0.35	002.06	0
002	Protein, Crude (%)	0011	12.151	0.05300	12.046	0.29606	0.11323	185	0.35	002.06	0
002	Protein, Crude (%)	0638	12.155	0.13000	12.046	0.29606	0.11323	185	0.37	002.04	0
002	Protein, Crude (%)	0148	12.155	0.01000	12.046	0.29606	0.11323	185	0.37	002.06	0
002	Protein, Crude (%)	0618	12.158	0.00400	12.046	0.29606	0.11323	185	0.38	002.06	0
002	Protein, Crude (%)	0682	12.170	0.00000	12.046	0.29606	0.11323	185	0.42	002.06	0
002	Protein, Crude (%)	2126	12.170	0.08000	12.046	0.29606	0.11323	185	0.42	002.06	0
002	Protein, Crude (%)	0019	12.175	0.01000	12.046	0.29606	0.11323	185	0.44	002.06	0
002	Protein, Crude (%)	0948	12.175	0.01000	12.046	0.29606	0.11323	185	0.44	002.06	0
002	Protein, Crude (%)	2075	12.180	0.06000	12.046	0.29606	0.11323	185	0.45	002.06	0
002	Protein, Crude (%)	0004	12.200	0.02000	12.046	0.29606	0.11323	185	0.52	002.06	0
002	Protein, Crude (%)	0010	12.200	0.00000	12.046	0.29606	0.11323	185	0.52	002.06	0
002	Protein, Crude (%)	0553	12.205	0.39000	12.046	0.29606	0.11323	185	0.54	002.06	0
002	Protein, Crude (%)	0298	12.210	0.02000	12.046	0.29606	0.11323	185	0.55	002.06	0
002	Protein, Crude (%)	0889	12.225	0.21000	12.046	0.29606	0.11323	185	0.60	002.01	0
002	Protein, Crude (%)	0722	12.238	0.06950	12.046	0.29606	0.11323	185	0.65	002.05	0
002	Protein, Crude (%)	0353	12.240	0.02000	12.046	0.29606	0.11323	185	0.66	002.06	0
002	Protein, Crude (%)	0152	12.250	0.04000	12.046	0.29606	0.11323	185	0.69	002.01	0
002	Protein, Crude (%)	0083	12.250	0.08000	12.046	0.29606	0.11323	185	0.69	002.06	0
002	Protein, Crude (%)	0510	12.250	0.10000	12.046	0.29606	0.11323	185	0.69	002.06	0
002	Protein, Crude (%)	0407	12.268	0.03600	12.046	0.29606	0.11323	185	0.75	002.06	0
002	Protein, Crude (%)	2128	12.282	0.06800	12.046	0.29606	0.11323	185	0.80	002.11	0
002	Protein, Crude (%)	0309	12.283	0.26850	12.046	0.29606	0.11323	185	0.80	002.06	0
002	Protein, Crude (%)	0037	12.290	0.30000	12.046	0.29606	0.11323	185	0.82	002.06	0
002	Protein, Crude (%)	0413	12.300	0.20000	12.046	0.29606	0.11323	185	0.86	002.06	0
002	Protein, Crude (%)	0574	12.305	0.01000	12.046	0.29606	0.11323	185	0.88	002.06	0
002	Protein, Crude (%)	0876	12.315	0.07000	12.046	0.29606	0.11323	185	0.91	002.06	0
002	Protein, Crude (%)	0028	12.330	0.28000	12.046	0.29606	0.11323	185	0.96	002.06	0
002	Protein, Crude (%)	0425	12.335	0.03000	12.046	0.29606	0.11323	185	0.98	002.06	0
002	Protein, Crude (%)	2193	12.340	0.02000	12.046	0.29606	0.11323	185	0.99	002.01	0
002	Protein, Crude (%)	0205	12.340	0.22000	12.046	0.29606	0.11323	185	0.99	002.06	0
002	Protein, Crude (%)	0229	12.342	0.21700	12.046	0.29606	0.11323	185	1.00	002.06	0
002	Protein, Crude (%)	0190	12.350	0.04000	12.046	0.29606	0.11323	185	1.03	002.06	0
002	Protein, Crude (%)	0884	12.350	0.10000	12.046	0.29606	0.11323	185	1.03	002.06	0
002	Protein, Crude (%)	0644	12.362	0.02400	12.046	0.29606	0.11323	185	1.07	002.06	0
002	Protein, Crude (%)	0033	12.400	0.40000	12.046	0.29606	0.11323	185	1.20	002.06	0
002	Protein, Crude (%)	0036	12.400	0.04000	12.046	0.29606	0.11323	185	1.20	002.06	0
002	Protein, Crude (%)	0596	12.405	0.15000	12.046	0.29606	0.11323	185	1.21	002.06	0
002	Protein, Crude (%)	0643	12.410	0.38000	12.046	0.29606	0.11323	185	1.23	002.99	0
002	Protein, Crude (%)	0358	12.415	0.19000	12.046	0.29606	0.11323	185	1.25	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein, Crude (%)	2006	12.435	0.17000	12.046	0.29606	0.11323	185	1.31	002.06	0
002	Protein, Crude (%)	0630	12.470	0.34000	12.046	0.29606	0.11323	185	1.43	002.06	0
002	Protein, Crude (%)	0511	12.480	0.12000	12.046	0.29606	0.11323	185	1.47	002.06	0
002	Protein, Crude (%)	0242	12.485	0.31000	12.046	0.29606	0.11323	185	1.48	002.06	0
002	Protein, Crude (%)	0843	12.485	0.03000	12.046	0.29606	0.11323	185	1.48	002.06	0
002	Protein, Crude (%)	0646	12.510	0.18000	12.046	0.29606	0.11323	185	1.57	002.06	0
002	Protein, Crude (%)	0035	12.550	0.26000	12.046	0.29606	0.11323	185	1.70	002.06	0
002	Protein, Crude (%)	0001	12.560	0.10000	12.046	0.29606	0.11323	185	1.74	002.06	0
002	Protein, Crude (%)	0186	12.600	0.20000	12.046	0.29606	0.11323	185	1.87	002.06	0
002	Protein, Crude (%)	0034	12.607	0.07900	12.046	0.29606	0.11323	185	1.89	002.06	0
002	Protein, Crude (%)	0553	12.665	0.09000	12.046	0.29606	0.11323	185	2.09	002.11	0
002	Protein, Crude (%)	2105	12.685	0.11000	12.046	0.29606	0.11323	185	2.16	002.11	0
002	Protein, Crude (%)	0233	12.705	0.37000	12.046	0.29606	0.11323	185	2.23	002.06	0
002	Protein, Crude (%)	2200	12.900	0.20000	12.046	0.29606	0.11323	185	2.88	002.05	0
002	Protein, Crude (%)	2146	13.155	0.03000	12.046	0.29606	0.11323	185	3.75	002.05	0
002	Protein, Crude (%)	0852	13.220	0.18000	12.046	0.29606	0.11323	185	3.97	002.05	0
002	Protein, Crude (%)	0089	13.950	0.00000	12.046	0.29606	0.11323	185	6.43	002.06	0
002	Protein, Crude (%)	0612	14.485	0.09000	12.046	0.29606	0.11323	185	8.24	002.06	0
002	Protein, Crude (%)	0011	15.100	0.00000	12.046	0.29606	0.11323	185	10.32	002.11	0
002	Protein, Crude (%)	0852	15.120	0.00000	12.046	0.29606	0.11323	185	10.38	002.11	0
002	Protein, Crude (%)	0889	9.3450	0.67000	12.046	0.29606	0.11323	185	-9.12	002.11	1
002	Protein, Crude (%)	0953	14.300	1.6400	12.046	0.29606	0.11323	185	7.61	002.06	1
003	Fat, Crude (%)	0357	0.80000	0.00000	2.2516	0.34080	0.10881	149	-4.26	003.14	0
003	Fat, Crude (%)	2129	1.1495	0.00100	2.2516	0.34080	0.10881	149	-3.23	003.14	0
003	Fat, Crude (%)	0880	1.1500	0.30000	2.2516	0.34080	0.10881	149	-3.23	003.99	0
003	Fat, Crude (%)	0889	1.2300	0.10000	2.2516	0.34080	0.10881	149	-3.00	003.11	0
003	Fat, Crude (%)	2105	1.4200	0.12000	2.2516	0.34080	0.10881	149	-2.44	003.11	0
003	Fat, Crude (%)	0358	1.4600	0.00000	2.2516	0.34080	0.10881	149	-2.32	003.14	0
003	Fat, Crude (%)	0042	1.5250	0.23000	2.2516	0.34080	0.10881	149	-2.13	003.10	0
003	Fat, Crude (%)	0175	1.5500	0.02000	2.2516	0.34080	0.10881	149	-2.06	003.14	0
003	Fat, Crude (%)	0581	1.6550	0.17000	2.2516	0.34080	0.10881	149	-1.75	003.14	0
003	Fat, Crude (%)	0897	1.7050	0.03000	2.2516	0.34080	0.10881	149	-1.60	003.10	0
003	Fat, Crude (%)	0019	1.7050	0.11000	2.2516	0.34080	0.10881	149	-1.60	003.14	0
003	Fat, Crude (%)	2155	1.7500	0.50000	2.2516	0.34080	0.10881	149	-1.47	003.10	0
003	Fat, Crude (%)	0034	1.7600	0.04000	2.2516	0.34080	0.10881	149	-1.44	003.14	0
003	Fat, Crude (%)	0142	1.7900	0.18000	2.2516	0.34080	0.10881	149	-1.35	003.00	0
003	Fat, Crude (%)	2181	1.7900	0.22000	2.2516	0.34080	0.10881	149	-1.35	003.06	0
003	Fat, Crude (%)	0037	1.7900	0.14000	2.2516	0.34080	0.10881	149	-1.35	003.14	0
003	Fat, Crude (%)	0144	1.8450	0.01000	2.2516	0.34080	0.10881	149	-1.19	003.14	0
003	Fat, Crude (%)	0298	1.8600	0.02000	2.2516	0.34080	0.10881	149	-1.15	003.10	0
003	Fat, Crude (%)	0876	1.8650	0.27000	2.2516	0.34080	0.10881	149	-1.13	003.14	0
003	Fat, Crude (%)	0868	1.8850	0.05000	2.2516	0.34080	0.10881	149	-1.08	003.14	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat, Crude (%)	2128	1.9000	0.14660	2.2516	0.34080	0.10881	149	-1.03	003.11	0
003	Fat, Crude (%)	0860	1.9200	0.00000	2.2516	0.34080	0.10881	149	-0.97	003.14	0
003	Fat, Crude (%)	0294	1.9250	0.09000	2.2516	0.34080	0.10881	149	-0.96	003.06	0
003	Fat, Crude (%)	0952	1.9500	0.08000	2.2516	0.34080	0.10881	149	-0.88	003.10	0
003	Fat, Crude (%)	0553	1.9550	0.07000	2.2516	0.34080	0.10881	149	-0.87	003.10	0
003	Fat, Crude (%)	0878	1.9550	0.01000	2.2516	0.34080	0.10881	149	-0.87	003.14	0
003	Fat, Crude (%)	0011	1.9650	0.07000	2.2516	0.34080	0.10881	149	-0.84	003.12	0
003	Fat, Crude (%)	0574	1.9650	0.07000	2.2516	0.34080	0.10881	149	-0.84	003.14	0
003	Fat, Crude (%)	0536	1.9850	0.05000	2.2516	0.34080	0.10881	149	-0.78	003.99	0
003	Fat, Crude (%)	0598	1.9950	0.25000	2.2516	0.34080	0.10881	149	-0.75	003.14	0
003	Fat, Crude (%)	0956	2.0000	0.00000	2.2516	0.34080	0.10881	149	-0.74	003.14	0
003	Fat, Crude (%)	0036	2.0271	0.03080	2.2516	0.34080	0.10881	149	-0.66	003.10	0
003	Fat, Crude (%)	0003	2.0400	0.24000	2.2516	0.34080	0.10881	149	-0.62	003.14	0
003	Fat, Crude (%)	0505	2.0400	0.16000	2.2516	0.34080	0.10881	149	-0.62	003.14	0
003	Fat, Crude (%)	0843	2.0400	0.18000	2.2516	0.34080	0.10881	149	-0.62	003.14	0
003	Fat, Crude (%)	0186	2.0500	0.14000	2.2516	0.34080	0.10881	149	-0.59	003.14	0
003	Fat, Crude (%)	0278	2.0500	0.10000	2.2516	0.34080	0.10881	149	-0.59	003.14	0
003	Fat, Crude (%)	0939	2.0600	0.08000	2.2516	0.34080	0.10881	149	-0.56	003.14	0
003	Fat, Crude (%)	0169	2.0650	0.01000	2.2516	0.34080	0.10881	149	-0.55	003.06	0
003	Fat, Crude (%)	0045	2.0650	0.01000	2.2516	0.34080	0.10881	149	-0.55	003.10	0
003	Fat, Crude (%)	0861	2.0800	0.06000	2.2516	0.34080	0.10881	149	-0.50	003.10	0
003	Fat, Crude (%)	2195	2.0850	0.05000	2.2516	0.34080	0.10881	149	-0.49	003.10	0
003	Fat, Crude (%)	2009	2.0883	0.00250	2.2516	0.34080	0.10881	149	-0.48	003.10	0
003	Fat, Crude (%)	0619	2.0900	0.22000	2.2516	0.34080	0.10881	149	-0.47	003.10	0
003	Fat, Crude (%)	0868	2.0900	0.00000	2.2516	0.34080	0.10881	149	-0.47	003.10	0
003	Fat, Crude (%)	0265	2.0900	0.12000	2.2516	0.34080	0.10881	149	-0.47	003.14	0
003	Fat, Crude (%)	0407	2.0929	0.17890	2.2516	0.34080	0.10881	149	-0.47	003.14	0
003	Fat, Crude (%)	0895	2.1050	0.01000	2.2516	0.34080	0.10881	149	-0.43	003.12	0
003	Fat, Crude (%)	0100	2.1100	0.16000	2.2516	0.34080	0.10881	149	-0.42	003.10	0
003	Fat, Crude (%)	0033	2.1100	0.16000	2.2516	0.34080	0.10881	149	-0.42	003.13	0
003	Fat, Crude (%)	2012	2.1150	0.05000	2.2516	0.34080	0.10881	149	-0.40	003.09	0
003	Fat, Crude (%)	0520	2.1150	0.05000	2.2516	0.34080	0.10881	149	-0.40	003.14	0
003	Fat, Crude (%)	0187	2.1200	0.02000	2.2516	0.34080	0.10881	149	-0.39	003.13	0
003	Fat, Crude (%)	0027	2.1245	0.00300	2.2516	0.34080	0.10881	149	-0.37	003.09	0
003	Fat, Crude (%)	0098	2.1300	0.16000	2.2516	0.34080	0.10881	149	-0.36	003.10	0
003	Fat, Crude (%)	0089	2.1350	0.01000	2.2516	0.34080	0.10881	149	-0.34	003.14	0
003	Fat, Crude (%)	0948	2.1400	0.02000	2.2516	0.34080	0.10881	149	-0.33	003.14	0
003	Fat, Crude (%)	0853	2.1500	0.10000	2.2516	0.34080	0.10881	149	-0.30	003.14	0
003	Fat, Crude (%)	2022	2.1550	0.01000	2.2516	0.34080	0.10881	149	-0.28	003.10	0
003	Fat, Crude (%)	0870	2.1604	0.01500	2.2516	0.34080	0.10881	149	-0.27	003.10	0
003	Fat, Crude (%)	0083	2.1650	0.07000	2.2516	0.34080	0.10881	149	-0.25	003.06	0
003	Fat, Crude (%)	0001	2.1700	0.12200	2.2516	0.34080	0.10881	149	-0.24	003.14	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat, Crude (%)	0881	2.1725	0.00300	2.2516	0.34080	0.10881	149	-0.23	003.10	0
003	Fat, Crude (%)	0638	2.1750	0.05000	2.2516	0.34080	0.10881	149	-0.22	003.09	0
003	Fat, Crude (%)	0202	2.1800	0.02000	2.2516	0.34080	0.10881	149	-0.21	003.14	0
003	Fat, Crude (%)	0205	2.1905	0.01500	2.2516	0.34080	0.10881	149	-0.18	003.13	0
003	Fat, Crude (%)	0098	2.1950	0.09000	2.2516	0.34080	0.10881	149	-0.17	003.13	0
003	Fat, Crude (%)	2146	2.2000	0.06000	2.2516	0.34080	0.10881	149	-0.15	003.14	0
003	Fat, Crude (%)	0541	2.2100	0.28000	2.2516	0.34080	0.10881	149	-0.12	003.10	0
003	Fat, Crude (%)	0208	2.2150	0.07000	2.2516	0.34080	0.10881	149	-0.11	003.13	0
003	Fat, Crude (%)	0504	2.2200	0.28000	2.2516	0.34080	0.10881	149	-0.09	003.01	0
003	Fat, Crude (%)	0405	2.2200	0.00000	2.2516	0.34080	0.10881	149	-0.09	003.11	0
003	Fat, Crude (%)	0682	2.2200	0.00000	2.2516	0.34080	0.10881	149	-0.09	003.14	0
003	Fat, Crude (%)	2192	2.2250	0.11000	2.2516	0.34080	0.10881	149	-0.08	003.00	0
003	Fat, Crude (%)	0004	2.2350	0.29000	2.2516	0.34080	0.10881	149	-0.05	003.09	0
003	Fat, Crude (%)	0345	2.2500	0.10000	2.2516	0.34080	0.10881	149	0.00	003.00	0
003	Fat, Crude (%)	0689	2.2500	0.10000	2.2516	0.34080	0.10881	149	0.00	003.06	0
003	Fat, Crude (%)	2109	2.2550	0.07000	2.2516	0.34080	0.10881	149	0.01	003.00	0
003	Fat, Crude (%)	2196	2.2600	0.00000	2.2516	0.34080	0.10881	149	0.02	003.10	0
003	Fat, Crude (%)	0693	2.2650	0.41000	2.2516	0.34080	0.10881	149	0.04	003.10	0
003	Fat, Crude (%)	0948	2.2700	0.02000	2.2516	0.34080	0.10881	149	0.05	003.09	0
003	Fat, Crude (%)	2190	2.2700	0.28000	2.2516	0.34080	0.10881	149	0.05	003.00	0
003	Fat, Crude (%)	0695	2.2750	0.00000	2.2516	0.34080	0.10881	149	0.07	003.10	0
003	Fat, Crude (%)	0066	2.2765	0.19700	2.2516	0.34080	0.10881	149	0.07	003.14	0
003	Fat, Crude (%)	0849	2.2780	0.08400	2.2516	0.34080	0.10881	149	0.08	003.06	0
003	Fat, Crude (%)	0674	2.2800	0.00000	2.2516	0.34080	0.10881	149	0.08	003.09	0
003	Fat, Crude (%)	2073	2.2850	0.33000	2.2516	0.34080	0.10881	149	0.10	003.10	0
003	Fat, Crude (%)	0229	2.2865	0.04900	2.2516	0.34080	0.10881	149	0.10	003.06	0
003	Fat, Crude (%)	0511	2.2900	0.18000	2.2516	0.34080	0.10881	149	0.11	003.06	0
003	Fat, Crude (%)	0004	2.2900	0.26000	2.2516	0.34080	0.10881	149	0.11	003.14	0
003	Fat, Crude (%)	0229	2.3030	0.06800	2.2516	0.34080	0.10881	149	0.15	003.14	0
003	Fat, Crude (%)	0049	2.3050	0.31000	2.2516	0.34080	0.10881	149	0.16	003.14	0
003	Fat, Crude (%)	0098	2.3050	0.17000	2.2516	0.34080	0.10881	149	0.16	003.09	0
003	Fat, Crude (%)	0675	2.3100	0.10000	2.2516	0.34080	0.10881	149	0.17	003.09	0
003	Fat, Crude (%)	2188	2.3100	0.02000	2.2516	0.34080	0.10881	149	0.17	003.10	0
003	Fat, Crude (%)	2053	2.3350	0.13000	2.2516	0.34080	0.10881	149	0.24	003.00	0
003	Fat, Crude (%)	0662	2.3400	0.04000	2.2516	0.34080	0.10881	149	0.26	003.99	0
003	Fat, Crude (%)	0242	2.3500	0.06000	2.2516	0.34080	0.10881	149	0.29	003.10	0
003	Fat, Crude (%)	2006	2.3550	0.11000	2.2516	0.34080	0.10881	149	0.30	003.13	0
003	Fat, Crude (%)	0619	2.3700	0.08000	2.2516	0.34080	0.10881	149	0.35	003.14	0
003	Fat, Crude (%)	0955	2.3700	0.10000	2.2516	0.34080	0.10881	149	0.35	003.14	0
003	Fat, Crude (%)	0918	2.3850	0.09000	2.2516	0.34080	0.10881	149	0.39	003.06	0
003	Fat, Crude (%)	0689	2.4000	0.00000	2.2516	0.34080	0.10881	149	0.44	003.10	0
003	Fat, Crude (%)	2193	2.4000	0.00000	2.2516	0.34080	0.10881	149	0.44	003.10	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat, Crude (%)	0644	2.4020	0.09800	2.2516	0.34080	0.10881	149	0.44	003.06	0
003	Fat, Crude (%)	0968	2.4020	0.11400	2.2516	0.34080	0.10881	149	0.44	003.14	0
003	Fat, Crude (%)	0074	2.4250	0.29000	2.2516	0.34080	0.10881	149	0.51	003.14	0
003	Fat, Crude (%)	0910	2.4350	0.27000	2.2516	0.34080	0.10881	149	0.54	003.06	0
003	Fat, Crude (%)	0171	2.4450	0.05000	2.2516	0.34080	0.10881	149	0.57	003.12	0
003	Fat, Crude (%)	0889	2.4500	0.38000	2.2516	0.34080	0.10881	149	0.58	003.10	0
003	Fat, Crude (%)	2200	2.4500	0.30000	2.2516	0.34080	0.10881	149	0.58	003.12	0
003	Fat, Crude (%)	0038	2.4700	0.14000	2.2516	0.34080	0.10881	149	0.64	003.09	0
003	Fat, Crude (%)	0870	2.4798	0.01820	2.2516	0.34080	0.10881	149	0.67	003.13	0
003	Fat, Crude (%)	0354	2.4950	0.03000	2.2516	0.34080	0.10881	149	0.71	003.09	0
003	Fat, Crude (%)	0941	2.4950	0.07000	2.2516	0.34080	0.10881	149	0.71	003.99	0
003	Fat, Crude (%)	0884	2.5000	0.20000	2.2516	0.34080	0.10881	149	0.73	003.06	0
003	Fat, Crude (%)	0148	2.5000	0.00000	2.2516	0.34080	0.10881	149	0.73	003.14	0
003	Fat, Crude (%)	0009	2.5200	0.12000	2.2516	0.34080	0.10881	149	0.79	003.14	0
003	Fat, Crude (%)	0630	2.5300	0.00000	2.2516	0.34080	0.10881	149	0.82	003.09	0
003	Fat, Crude (%)	0108	2.5300	0.40000	2.2516	0.34080	0.10881	149	0.82	003.14	0
003	Fat, Crude (%)	0651	2.5570	0.00400	2.2516	0.34080	0.10881	149	0.90	003.10	0
003	Fat, Crude (%)	0175	2.5600	0.22000	2.2516	0.34080	0.10881	149	0.90	003.00	0
003	Fat, Crude (%)	0123	2.5650	0.01000	2.2516	0.34080	0.10881	149	0.92	003.13	0
003	Fat, Crude (%)	0596	2.5850	0.01000	2.2516	0.34080	0.10881	149	0.98	003.00	0
003	Fat, Crude (%)	0226	2.6000	0.20000	2.2516	0.34080	0.10881	149	1.02	003.09	0
003	Fat, Crude (%)	0685	2.6050	0.09000	2.2516	0.34080	0.10881	149	1.04	003.09	0
003	Fat, Crude (%)	2075	2.6150	0.11000	2.2516	0.34080	0.10881	149	1.07	003.14	0
003	Fat, Crude (%)	0164	2.6350	0.11000	2.2516	0.34080	0.10881	149	1.13	003.06	0
003	Fat, Crude (%)	0653	2.6450	0.07000	2.2516	0.34080	0.10881	149	1.15	003.00	0
003	Fat, Crude (%)	0413	2.6500	0.30000	2.2516	0.34080	0.10881	149	1.17	003.14	0
003	Fat, Crude (%)	0646	2.6600	0.02000	2.2516	0.34080	0.10881	149	1.20	003.13	0
003	Fat, Crude (%)	2126	2.6850	0.15000	2.2516	0.34080	0.10881	149	1.27	003.00	0
003	Fat, Crude (%)	0353	2.7700	0.08000	2.2516	0.34080	0.10881	149	1.52	003.09	0
003	Fat, Crude (%)	0011	2.8000	0.00000	2.2516	0.34080	0.10881	149	1.61	003.11	0
003	Fat, Crude (%)	0233	2.8200	0.12000	2.2516	0.34080	0.10881	149	1.67	003.10	0
003	Fat, Crude (%)	0152	2.8200	0.24000	2.2516	0.34080	0.10881	149	1.67	003.00	0
003	Fat, Crude (%)	0687	2.8200	0.36000	2.2516	0.34080	0.10881	149	1.67	003.06	0
003	Fat, Crude (%)	0190	2.8200	0.10000	2.2516	0.34080	0.10881	149	1.67	003.14	0
003	Fat, Crude (%)	0722	2.8410	0.02890	2.2516	0.34080	0.10881	149	1.73	003.09	0
003	Fat, Crude (%)	0017	2.8500	0.10000	2.2516	0.34080	0.10881	149	1.76	003.00	0
003	Fat, Crude (%)	0618	2.8950	0.11200	2.2516	0.34080	0.10881	149	1.89	003.10	0
003	Fat, Crude (%)	0510	2.9000	0.00000	2.2516	0.34080	0.10881	149	1.90	003.09	0
003	Fat, Crude (%)	2199	2.9500	0.30000	2.2516	0.34080	0.10881	149	2.05	003.12	0
003	Fat, Crude (%)	0870	2.9657	0.16330	2.2516	0.34080	0.10881	149	2.10	003.09	0
003	Fat, Crude (%)	0682	3.0000	0.00000	2.2516	0.34080	0.10881	149	2.20	003.06	0
003	Fat, Crude (%)	0309	3.0225	0.18890	2.2516	0.34080	0.10881	149	2.26	003.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat, Crude (%)	2144	3.0650	0.03000	2.2516	0.34080	0.10881	149	2.39	003.99	0
003	Fat, Crude (%)	0852	3.4300	0.04000	2.2516	0.34080	0.10881	149	3.46	003.11	0
003	Fat, Crude (%)	0676	3.9200	0.06000	2.2516	0.34080	0.10881	149	4.90	003.10	0
003	Fat, Crude (%)	2089	1.5350	0.75000	2.2516	0.34080	0.10881	149	-2.10	003.13	1
004	Fiber, Crude (%)	2089	13.845	0.19000	17.899	1.2825	0.35456	116	-3.16	004.03	0
004	Fiber, Crude (%)	2146	15.680	0.56000	17.899	1.2825	0.35456	116	-1.73	004.99	0
004	Fiber, Crude (%)	2012	15.740	0.14000	17.899	1.2825	0.35456	116	-1.68	004.07	0
004	Fiber, Crude (%)	0510	15.750	0.10000	17.899	1.2825	0.35456	116	-1.68	004.07	0
004	Fiber, Crude (%)	0505	15.790	0.14000	17.899	1.2825	0.35456	116	-1.64	004.07	0
004	Fiber, Crude (%)	0265	15.800	0.80000	17.899	1.2825	0.35456	116	-1.64	004.07	0
004	Fiber, Crude (%)	0682	15.800	0.00000	17.899	1.2825	0.35456	116	-1.64	004.07	0
004	Fiber, Crude (%)	2145	15.975	0.11000	17.899	1.2825	0.35456	116	-1.50	004.99	0
004	Fiber, Crude (%)	0123	16.020	0.08000	17.899	1.2825	0.35456	116	-1.47	004.07	0
004	Fiber, Crude (%)	2193	16.115	0.03000	17.899	1.2825	0.35456	116	-1.39	004.07	0
004	Fiber, Crude (%)	2109	16.220	0.64000	17.899	1.2825	0.35456	116	-1.31	004.07	0
004	Fiber, Crude (%)	0004	16.280	0.90000	17.899	1.2825	0.35456	116	-1.26	004.07	0
004	Fiber, Crude (%)	0098	16.320	0.86000	17.899	1.2825	0.35456	116	-1.23	004.07	0
004	Fiber, Crude (%)	0897	16.410	0.02000	17.899	1.2825	0.35456	116	-1.16	004.06	0
004	Fiber, Crude (%)	0407	16.518	0.29380	17.899	1.2825	0.35456	116	-1.08	004.07	0
004	Fiber, Crude (%)	2006	16.540	0.00000	17.899	1.2825	0.35456	116	-1.06	004.06	0
004	Fiber, Crude (%)	0413	16.650	0.70000	17.899	1.2825	0.35456	116	-0.97	004.07	0
004	Fiber, Crude (%)	0028	16.700	0.40000	17.899	1.2825	0.35456	116	-0.94	004.07	0
004	Fiber, Crude (%)	0202	16.725	0.11000	17.899	1.2825	0.35456	116	-0.92	004.07	0
004	Fiber, Crude (%)	0035	16.730	0.98000	17.899	1.2825	0.35456	116	-0.91	004.07	0
004	Fiber, Crude (%)	0870	16.757	0.01330	17.899	1.2825	0.35456	116	-0.89	004.07	0
004	Fiber, Crude (%)	2144	16.795	0.13000	17.899	1.2825	0.35456	116	-0.86	004.99	0
004	Fiber, Crude (%)	0098	16.810	0.44000	17.899	1.2825	0.35456	116	-0.85	004.06	0
004	Fiber, Crude (%)	0610	16.850	0.50000	17.899	1.2825	0.35456	116	-0.82	004.99	0
004	Fiber, Crude (%)	0619	16.900	0.40000	17.899	1.2825	0.35456	116	-0.78	004.07	0
004	Fiber, Crude (%)	0011	16.900	0.00000	17.899	1.2825	0.35456	116	-0.78	004.11	0
004	Fiber, Crude (%)	0405	17.020	0.00000	17.899	1.2825	0.35456	116	-0.69	004.11	0
004	Fiber, Crude (%)	2190	17.175	0.39000	17.899	1.2825	0.35456	116	-0.56	004.03	0
004	Fiber, Crude (%)	0205	17.200	0.00000	17.899	1.2825	0.35456	116	-0.55	004.06	0
004	Fiber, Crude (%)	0689	17.200	0.20000	17.899	1.2825	0.35456	116	-0.55	004.06	0
004	Fiber, Crude (%)	0895	17.235	0.01000	17.899	1.2825	0.35456	116	-0.52	004.06	0
004	Fiber, Crude (%)	0298	17.300	0.02000	17.899	1.2825	0.35456	116	-0.47	004.00	0
004	Fiber, Crude (%)	0939	17.300	0.46000	17.899	1.2825	0.35456	116	-0.47	004.07	0
004	Fiber, Crude (%)	0354	17.305	0.35000	17.899	1.2825	0.35456	116	-0.46	004.00	0
004	Fiber, Crude (%)	0353	17.320	0.08000	17.899	1.2825	0.35456	116	-0.45	004.03	0
004	Fiber, Crude (%)	2188	17.325	0.09000	17.899	1.2825	0.35456	116	-0.45	004.07	0
004	Fiber, Crude (%)	0345	17.350	0.50000	17.899	1.2825	0.35456	116	-0.43	004.00	0
004	Fiber, Crude (%)	0190	17.350	0.10000	17.899	1.2825	0.35456	116	-0.43	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber, Crude (%)	0860	17.370	0.00000	17.899	1.2825	0.35456	116	-0.41	004.06	0
004	Fiber, Crude (%)	0884	17.400	0.40000	17.899	1.2825	0.35456	116	-0.39	004.00	0
004	Fiber, Crude (%)	0229	17.400	0.50000	17.899	1.2825	0.35456	116	-0.39	004.07	0
004	Fiber, Crude (%)	0722	17.410	0.06940	17.899	1.2825	0.35456	116	-0.38	004.06	0
004	Fiber, Crude (%)	2126	17.445	0.15000	17.899	1.2825	0.35456	116	-0.35	004.99	0
004	Fiber, Crude (%)	0354	17.450	0.02000	17.899	1.2825	0.35456	116	-0.35	004.07	0
004	Fiber, Crude (%)	0100	17.465	0.75000	17.899	1.2825	0.35456	116	-0.34	004.07	0
004	Fiber, Crude (%)	0590	17.475	0.39000	17.899	1.2825	0.35456	116	-0.33	004.06	0
004	Fiber, Crude (%)	0918	17.485	0.27000	17.899	1.2825	0.35456	116	-0.32	004.06	0
004	Fiber, Crude (%)	2190	17.490	0.06000	17.899	1.2825	0.35456	116	-0.32	004.07	0
004	Fiber, Crude (%)	0175	17.550	0.10000	17.899	1.2825	0.35456	116	-0.27	004.00	0
004	Fiber, Crude (%)	2022	17.550	0.14000	17.899	1.2825	0.35456	116	-0.27	004.06	0
004	Fiber, Crude (%)	0644	17.632	1.2880	17.899	1.2825	0.35456	116	-0.21	004.07	0
004	Fiber, Crude (%)	0027	17.680	0.06100	17.899	1.2825	0.35456	116	-0.17	004.06	0
004	Fiber, Crude (%)	0881	17.745	0.03900	17.899	1.2825	0.35456	116	-0.12	004.06	0
004	Fiber, Crude (%)	0968	17.817	0.09600	17.899	1.2825	0.35456	116	-0.06	004.06	0
004	Fiber, Crude (%)	2146	17.820	0.54000	17.899	1.2825	0.35456	116	-0.06	004.07	0
004	Fiber, Crude (%)	0849	17.850	0.70000	17.899	1.2825	0.35456	116	-0.04	004.07	0
004	Fiber, Crude (%)	2181	17.855	0.23000	17.899	1.2825	0.35456	116	-0.03	004.07	0
004	Fiber, Crude (%)	0169	17.860	0.68000	17.899	1.2825	0.35456	116	-0.03	004.00	0
004	Fiber, Crude (%)	2009	17.893	0.02490	17.899	1.2825	0.35456	116	0.00	004.07	0
004	Fiber, Crude (%)	0309	17.896	0.43600	17.899	1.2825	0.35456	116	0.00	004.00	0
004	Fiber, Crude (%)	0968	17.909	0.05100	17.899	1.2825	0.35456	116	0.01	004.07	0
004	Fiber, Crude (%)	0948	17.920	0.36000	17.899	1.2825	0.35456	116	0.02	004.07	0
004	Fiber, Crude (%)	0171	17.970	1.2600	17.899	1.2825	0.35456	116	0.06	004.00	0
004	Fiber, Crude (%)	0948	17.975	0.29000	17.899	1.2825	0.35456	116	0.06	004.06	0
004	Fiber, Crude (%)	0425	18.000	0.20000	17.899	1.2825	0.35456	116	0.08	004.00	0
004	Fiber, Crude (%)	0952	18.080	0.06000	17.899	1.2825	0.35456	116	0.14	004.06	0
004	Fiber, Crude (%)	0581	18.085	0.11000	17.899	1.2825	0.35456	116	0.14	004.07	0
004	Fiber, Crude (%)	0596	18.120	1.0000	17.899	1.2825	0.35456	116	0.17	004.07	0
004	Fiber, Crude (%)	0868	18.130	0.00000	17.899	1.2825	0.35456	116	0.18	004.06	0
004	Fiber, Crude (%)	2073	18.140	0.18000	17.899	1.2825	0.35456	116	0.19	004.00	0
004	Fiber, Crude (%)	2192	18.155	0.99000	17.899	1.2825	0.35456	116	0.20	004.03	0
004	Fiber, Crude (%)	2128	18.177	0.70900	17.899	1.2825	0.35456	116	0.22	004.11	0
004	Fiber, Crude (%)	0910	18.200	0.40000	17.899	1.2825	0.35456	116	0.23	004.07	0
004	Fiber, Crude (%)	0878	18.235	0.37000	17.899	1.2825	0.35456	116	0.26	004.06	0
004	Fiber, Crude (%)	0278	18.250	0.10000	17.899	1.2825	0.35456	116	0.27	004.07	0
004	Fiber, Crude (%)	0674	18.270	0.00000	17.899	1.2825	0.35456	116	0.29	004.06	0
004	Fiber, Crude (%)	2196	18.290	0.00000	17.899	1.2825	0.35456	116	0.30	004.07	0
004	Fiber, Crude (%)	0009	18.320	0.44000	17.899	1.2825	0.35456	116	0.33	004.07	0
004	Fiber, Crude (%)	0956	18.350	0.50000	17.899	1.2825	0.35456	116	0.35	004.07	0
004	Fiber, Crude (%)	0164	18.400	0.20000	17.899	1.2825	0.35456	116	0.39	004.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber, Crude (%)	0208	18.400	0.00000	17.899	1.2825	0.35456	116	0.39	004.00	0
004	Fiber, Crude (%)	0074	18.440	0.76000	17.899	1.2825	0.35456	116	0.42	004.07	0
004	Fiber, Crude (%)	0941	18.450	0.30000	17.899	1.2825	0.35456	116	0.43	004.06	0
004	Fiber, Crude (%)	2200	18.500	0.40000	17.899	1.2825	0.35456	116	0.47	004.06	0
004	Fiber, Crude (%)	0675	18.535	0.27000	17.899	1.2825	0.35456	116	0.50	004.06	0
004	Fiber, Crude (%)	0876	18.595	0.01000	17.899	1.2825	0.35456	116	0.54	004.07	0
004	Fiber, Crude (%)	0955	18.650	1.1000	17.899	1.2825	0.35456	116	0.59	004.07	0
004	Fiber, Crude (%)	0294	18.700	0.00000	17.899	1.2825	0.35456	116	0.62	004.07	0
004	Fiber, Crude (%)	0598	18.715	0.89000	17.899	1.2825	0.35456	116	0.64	004.07	0
004	Fiber, Crude (%)	0034	18.770	0.92000	17.899	1.2825	0.35456	116	0.68	004.07	0
004	Fiber, Crude (%)	0536	18.805	0.13000	17.899	1.2825	0.35456	116	0.71	004.07	0
004	Fiber, Crude (%)	0653	18.840	0.38000	17.899	1.2825	0.35456	116	0.73	004.07	0
004	Fiber, Crude (%)	0852	18.855	0.31000	17.899	1.2825	0.35456	116	0.75	004.11	0
004	Fiber, Crude (%)	0066	18.856	0.30300	17.899	1.2825	0.35456	116	0.75	004.07	0
004	Fiber, Crude (%)	0553	18.910	0.74000	17.899	1.2825	0.35456	116	0.79	004.07	0
004	Fiber, Crude (%)	0083	19.030	0.42000	17.899	1.2825	0.35456	116	0.88	004.07	0
004	Fiber, Crude (%)	2053	19.155	0.05000	17.899	1.2825	0.35456	116	0.98	004.07	0
004	Fiber, Crude (%)	0708	19.185	0.07000	17.899	1.2825	0.35456	116	1.00	004.07	0
004	Fiber, Crude (%)	0511	19.250	0.50000	17.899	1.2825	0.35456	116	1.05	004.00	0
004	Fiber, Crude (%)	0861	19.395	0.27000	17.899	1.2825	0.35456	116	1.17	004.07	0
004	Fiber, Crude (%)	2129	19.635	0.49000	17.899	1.2825	0.35456	116	1.35	004.99	0
004	Fiber, Crude (%)	0019	19.670	0.88000	17.899	1.2825	0.35456	116	1.38	004.07	0
004	Fiber, Crude (%)	0643	19.720	0.24000	17.899	1.2825	0.35456	116	1.42	004.07	0
004	Fiber, Crude (%)	0089	19.730	0.24000	17.899	1.2825	0.35456	116	1.43	004.07	0
004	Fiber, Crude (%)	0693	19.917	0.26500	17.899	1.2825	0.35456	116	1.57	004.03	0
004	Fiber, Crude (%)	0032	20.080	1.4000	17.899	1.2825	0.35456	116	1.70	004.07	0
004	Fiber, Crude (%)	0226	20.185	0.17000	17.899	1.2825	0.35456	116	1.78	004.00	0
004	Fiber, Crude (%)	0646	20.360	0.10000	17.899	1.2825	0.35456	116	1.92	004.07	0
004	Fiber, Crude (%)	0242	20.450	0.18000	17.899	1.2825	0.35456	116	1.99	004.07	0
004	Fiber, Crude (%)	0144	20.570	0.30000	17.899	1.2825	0.35456	116	2.08	004.07	0
004	Fiber, Crude (%)	0045	20.750	0.50000	17.899	1.2825	0.35456	116	2.22	004.07	0
004	Fiber, Crude (%)	0520	21.350	1.7000	17.899	1.2825	0.35456	116	2.69	004.07	0
004	Fiber, Crude (%)	0504	21.605	1.1900	17.899	1.2825	0.35456	116	2.89	004.00	0
004	Fiber, Crude (%)	0626	21.670	0.02000	17.899	1.2825	0.35456	116	2.94	004.07	0
004	Fiber, Crude (%)	2105	23.850	0.38000	17.899	1.2825	0.35456	116	4.64	004.11	0
004	Fiber, Crude (%)	2195	28.305	0.35000	17.899	1.2825	0.35456	116	8.11	004.07	0
004	Fiber, Crude (%)	0003	19.060	1.9600	17.899	1.2825	0.35456	116	0.91	004.07	1
004	Fiber, Crude (%)	0638	19.150	2.1000	17.899	1.2825	0.35456	116	0.98	004.06	1
004	Fiber, Crude (%)	0676	0.62000	0.12000	17.899	1.2825	0.35456	116	-13.47	004.06	2
005	Ash (%)	0852	10.195	0.01000	16.957	0.63990	0.17516	139	-10.57	005.11	0
005	Ash (%)	0171	10.240	0.02000	16.957	0.63990	0.17516	139	-10.50	005.00	0
005	Ash (%)	2128	10.385	0.34400	16.957	0.63990	0.17516	139	-10.27	005.11	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	2105	12.725	0.41000	16.957	0.63990	0.17516	139	-6.61	005.11	0
005	Ash (%)	2081	14.555	0.31000	16.957	0.63990	0.17516	139	-3.75	005.99	0
005	Ash (%)	0164	15.400	0.40000	16.957	0.63990	0.17516	139	-2.43	005.00	0
005	Ash (%)	0964	15.871	0.22000	16.957	0.63990	0.17516	139	-1.70	005.00	0
005	Ash (%)	0169	15.910	0.10000	16.957	0.63990	0.17516	139	-1.64	005.00	0
005	Ash (%)	2200	15.950	0.50000	16.957	0.63990	0.17516	139	-1.57	005.00	0
005	Ash (%)	0511	16.005	0.31000	16.957	0.63990	0.17516	139	-1.49	005.00	0
005	Ash (%)	0278	16.020	0.02000	16.957	0.63990	0.17516	139	-1.46	005.00	0
005	Ash (%)	0357	16.100	0.20000	16.957	0.63990	0.17516	139	-1.34	005.00	0
005	Ash (%)	0662	16.129	0.72310	16.957	0.63990	0.17516	139	-1.29	005.99	0
005	Ash (%)	0650	16.195	0.27000	16.957	0.63990	0.17516	139	-1.19	005.00	0
005	Ash (%)	0589	16.215	0.45000	16.957	0.63990	0.17516	139	-1.16	005.00	0
005	Ash (%)	0622	16.229	0.01330	16.957	0.63990	0.17516	139	-1.14	005.00	0
005	Ash (%)	0425	16.250	0.04000	16.957	0.63990	0.17516	139	-1.11	005.00	0
005	Ash (%)	0208	16.300	0.00000	16.957	0.63990	0.17516	139	-1.03	005.00	0
005	Ash (%)	2066	16.300	0.40000	16.957	0.63990	0.17516	139	-1.03	005.00	0
005	Ash (%)	2073	16.305	0.01000	16.957	0.63990	0.17516	139	-1.02	005.00	0
005	Ash (%)	0034	16.306	0.71800	16.957	0.63990	0.17516	139	-1.02	005.00	0
005	Ash (%)	0539	16.335	0.15000	16.957	0.63990	0.17516	139	-0.97	005.00	0
005	Ash (%)	0574	16.335	0.11000	16.957	0.63990	0.17516	139	-0.97	005.99	0
005	Ash (%)	0148	16.345	0.25000	16.957	0.63990	0.17516	139	-0.96	005.00	0
005	Ash (%)	0019	16.400	0.60000	16.957	0.63990	0.17516	139	-0.87	005.00	0
005	Ash (%)	0003	16.400	0.00000	16.957	0.63990	0.17516	139	-0.87	005.05	0
005	Ash (%)	0722	16.405	0.00150	16.957	0.63990	0.17516	139	-0.86	005.00	0
005	Ash (%)	0682	16.440	0.00000	16.957	0.63990	0.17516	139	-0.81	005.00	0
005	Ash (%)	0175	16.450	0.50000	16.957	0.63990	0.17516	139	-0.79	005.00	0
005	Ash (%)	0186	16.450	0.30000	16.957	0.63990	0.17516	139	-0.79	005.00	0
005	Ash (%)	0598	16.535	0.05000	16.957	0.63990	0.17516	139	-0.66	005.00	0
005	Ash (%)	0541	16.540	0.22000	16.957	0.63990	0.17516	139	-0.65	005.00	0
005	Ash (%)	2181	16.540	0.26000	16.957	0.63990	0.17516	139	-0.65	005.00	0
005	Ash (%)	2199	16.550	0.10000	16.957	0.63990	0.17516	139	-0.64	005.00	0
005	Ash (%)	0596	16.555	0.19000	16.957	0.63990	0.17516	139	-0.63	005.00	0
005	Ash (%)	0651	16.555	0.01800	16.957	0.63990	0.17516	139	-0.63	005.00	0
005	Ash (%)	0618	16.569	0.12100	16.957	0.63990	0.17516	139	-0.61	005.00	0
005	Ash (%)	0066	16.579	0.11800	16.957	0.63990	0.17516	139	-0.59	005.05	0
005	Ash (%)	0520	16.585	0.21000	16.957	0.63990	0.17516	139	-0.58	005.00	0
005	Ash (%)	0004	16.590	0.28000	16.957	0.63990	0.17516	139	-0.57	005.00	0
005	Ash (%)	0693	16.595	0.35100	16.957	0.63990	0.17516	139	-0.57	005.00	0
005	Ash (%)	0881	16.609	0.01000	16.957	0.63990	0.17516	139	-0.54	005.00	0
005	Ash (%)	0918	16.625	0.09000	16.957	0.63990	0.17516	139	-0.52	005.00	0
005	Ash (%)	0038	16.630	0.38000	16.957	0.63990	0.17516	139	-0.51	005.05	0
005	Ash (%)	2188	16.635	0.21000	16.957	0.63990	0.17516	139	-0.50	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	2022	16.645	0.01000	16.957	0.63990	0.17516	139	-0.49	005.00	0
005	Ash (%)	0876	16.660	0.06000	16.957	0.63990	0.17516	139	-0.46	005.00	0
005	Ash (%)	2053	16.665	0.19000	16.957	0.63990	0.17516	139	-0.46	005.00	0
005	Ash (%)	0108	16.670	0.52000	16.957	0.63990	0.17516	139	-0.45	005.00	0
005	Ash (%)	0098	16.675	0.27000	16.957	0.63990	0.17516	139	-0.44	005.00	0
005	Ash (%)	0142	16.680	0.04000	16.957	0.63990	0.17516	139	-0.43	005.00	0
005	Ash (%)	0049	16.690	0.50000	16.957	0.63990	0.17516	139	-0.42	005.00	0
005	Ash (%)	0510	16.705	0.15000	16.957	0.63990	0.17516	139	-0.39	005.00	0
005	Ash (%)	0675	16.710	0.08000	16.957	0.63990	0.17516	139	-0.39	005.00	0
005	Ash (%)	0870	16.726	0.09390	16.957	0.63990	0.17516	139	-0.36	005.00	0
005	Ash (%)	0861	16.740	0.18000	16.957	0.63990	0.17516	139	-0.34	005.00	0
005	Ash (%)	0956	16.750	0.50000	16.957	0.63990	0.17516	139	-0.32	005.00	0
005	Ash (%)	0205	16.760	0.02000	16.957	0.63990	0.17516	139	-0.31	005.00	0
005	Ash (%)	0953	16.770	0.20000	16.957	0.63990	0.17516	139	-0.29	005.00	0
005	Ash (%)	2150	16.780	0.22000	16.957	0.63990	0.17516	139	-0.28	005.00	0
005	Ash (%)	2012	16.785	0.09000	16.957	0.63990	0.17516	139	-0.27	005.00	0
005	Ash (%)	0027	16.812	0.23400	16.957	0.63990	0.17516	139	-0.23	005.00	0
005	Ash (%)	0042	16.820	0.22000	16.957	0.63990	0.17516	139	-0.21	005.00	0
005	Ash (%)	0100	16.845	0.21000	16.957	0.63990	0.17516	139	-0.18	005.00	0
005	Ash (%)	0033	16.850	0.10000	16.957	0.63990	0.17516	139	-0.17	005.00	0
005	Ash (%)	2089	16.860	0.22000	16.957	0.63990	0.17516	139	-0.15	005.00	0
005	Ash (%)	0358	16.890	0.14000	16.957	0.63990	0.17516	139	-0.10	005.00	0
005	Ash (%)	0653	16.925	0.03000	16.957	0.63990	0.17516	139	-0.05	005.00	0
005	Ash (%)	0229	16.930	0.28000	16.957	0.63990	0.17516	139	-0.04	005.00	0
005	Ash (%)	0638	16.975	0.05000	16.957	0.63990	0.17516	139	0.03	005.05	0
005	Ash (%)	0035	16.980	0.10000	16.957	0.63990	0.17516	139	0.04	005.00	0
005	Ash (%)	0504	17.025	0.39000	16.957	0.63990	0.17516	139	0.11	005.00	0
005	Ash (%)	0646	17.030	0.32000	16.957	0.63990	0.17516	139	0.11	005.00	0
005	Ash (%)	2193	17.030	0.12000	16.957	0.63990	0.17516	139	0.11	005.00	0
005	Ash (%)	0689	17.045	0.01000	16.957	0.63990	0.17516	139	0.14	005.05	0
005	Ash (%)	0001	17.053	0.03370	16.957	0.63990	0.17516	139	0.15	005.00	0
005	Ash (%)	0874	17.060	0.36000	16.957	0.63990	0.17516	139	0.16	005.00	0
005	Ash (%)	0354	17.095	0.05000	16.957	0.63990	0.17516	139	0.22	005.00	0
005	Ash (%)	0505	17.095	0.01000	16.957	0.63990	0.17516	139	0.22	005.00	0
005	Ash (%)	0298	17.100	0.02000	16.957	0.63990	0.17516	139	0.22	005.00	0
005	Ash (%)	0880	17.100	0.20000	16.957	0.63990	0.17516	139	0.22	005.00	0
005	Ash (%)	0619	17.100	0.20000	16.957	0.63990	0.17516	139	0.22	005.05	0
005	Ash (%)	2006	17.110	0.04000	16.957	0.63990	0.17516	139	0.24	005.00	0
005	Ash (%)	0536	17.125	0.05000	16.957	0.63990	0.17516	139	0.26	005.99	0
005	Ash (%)	2196	17.130	0.00000	16.957	0.63990	0.17516	139	0.27	005.00	0
005	Ash (%)	2195	17.130	0.02000	16.957	0.63990	0.17516	139	0.27	005.00	0
005	Ash (%)	0407	17.140	0.03000	16.957	0.63990	0.17516	139	0.29	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0265	17.150	0.10000	16.957	0.63990	0.17516	139	0.30	005.00	0
005	Ash (%)	2109	17.150	0.30000	16.957	0.63990	0.17516	139	0.30	005.05	0
005	Ash (%)	0553	17.155	0.15000	16.957	0.63990	0.17516	139	0.31	005.00	0
005	Ash (%)	0674	17.160	0.00000	16.957	0.63990	0.17516	139	0.32	005.00	0
005	Ash (%)	0968	17.165	0.37900	16.957	0.63990	0.17516	139	0.32	005.00	0
005	Ash (%)	0948	17.170	0.04000	16.957	0.63990	0.17516	139	0.33	005.00	0
005	Ash (%)	0083	17.180	0.12000	16.957	0.63990	0.17516	139	0.35	005.00	0
005	Ash (%)	0144	17.185	0.01000	16.957	0.63990	0.17516	139	0.36	005.00	0
005	Ash (%)	0045	17.200	0.40000	16.957	0.63990	0.17516	139	0.38	005.00	0
005	Ash (%)	0853	17.200	0.40000	16.957	0.63990	0.17516	139	0.38	005.00	0
005	Ash (%)	0047	17.200	0.80000	16.957	0.63990	0.17516	139	0.38	005.00	0
005	Ash (%)	0353	17.210	0.16000	16.957	0.63990	0.17516	139	0.40	005.00	0
005	Ash (%)	0590	17.260	0.14000	16.957	0.63990	0.17516	139	0.47	005.05	0
005	Ash (%)	0581	17.300	0.06000	16.957	0.63990	0.17516	139	0.54	005.00	0
005	Ash (%)	2190	17.345	0.01000	16.957	0.63990	0.17516	139	0.61	005.05	0
005	Ash (%)	0849	17.360	0.02000	16.957	0.63990	0.17516	139	0.63	005.05	0
005	Ash (%)	0152	17.385	0.01000	16.957	0.63990	0.17516	139	0.67	005.05	0
005	Ash (%)	0190	17.400	0.20000	16.957	0.63990	0.17516	139	0.69	005.05	0
005	Ash (%)	0878	17.410	0.02000	16.957	0.63990	0.17516	139	0.71	005.05	0
005	Ash (%)	0242	17.530	0.18000	16.957	0.63990	0.17516	139	0.90	005.00	0
005	Ash (%)	0939	17.540	0.18000	16.957	0.63990	0.17516	139	0.91	005.99	0
005	Ash (%)	0895	17.545	0.01000	16.957	0.63990	0.17516	139	0.92	005.05	0
005	Ash (%)	0226	17.550	0.10000	16.957	0.63990	0.17516	139	0.93	005.00	0
005	Ash (%)	0202	17.565	0.25000	16.957	0.63990	0.17516	139	0.95	005.99	0
005	Ash (%)	0630	17.570	0.18000	16.957	0.63990	0.17516	139	0.96	005.00	0
005	Ash (%)	2145	17.595	0.03000	16.957	0.63990	0.17516	139	1.00	005.05	0
005	Ash (%)	0123	17.600	0.00000	16.957	0.63990	0.17516	139	1.00	005.00	0
005	Ash (%)	0897	17.615	0.05000	16.957	0.63990	0.17516	139	1.03	005.05	0
005	Ash (%)	2009	17.629	0.01390	16.957	0.63990	0.17516	139	1.05	005.05	0
005	Ash (%)	0643	17.635	0.33000	16.957	0.63990	0.17516	139	1.06	005.00	0
005	Ash (%)	0345	17.650	0.10000	16.957	0.63990	0.17516	139	1.08	005.00	0
005	Ash (%)	0139	17.695	0.17000	16.957	0.63990	0.17516	139	1.15	005.00	0
005	Ash (%)	0644	17.695	0.15000	16.957	0.63990	0.17516	139	1.15	005.02	0
005	Ash (%)	0413	17.700	0.80000	16.957	0.63990	0.17516	139	1.16	005.05	0
005	Ash (%)	0695	17.702	0.16300	16.957	0.63990	0.17516	139	1.16	005.05	0
005	Ash (%)	2146	17.710	0.28000	16.957	0.63990	0.17516	139	1.18	005.05	0
005	Ash (%)	0610	17.725	0.11000	16.957	0.63990	0.17516	139	1.20	005.05	0
005	Ash (%)	2097	17.738	0.03530	16.957	0.63990	0.17516	139	1.22	005.00	0
005	Ash (%)	0309	17.742	0.16600	16.957	0.63990	0.17516	139	1.23	005.00	0
005	Ash (%)	0294	17.750	0.10000	16.957	0.63990	0.17516	139	1.24	005.05	0
005	Ash (%)	0868	17.755	0.01000	16.957	0.63990	0.17516	139	1.25	005.05	0
005	Ash (%)	0884	17.850	0.10000	16.957	0.63990	0.17516	139	1.40	005.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	2192	17.880	0.06000	16.957	0.63990	0.17516	139	1.44	005.05	0
005	Ash (%)	0676	17.905	0.11000	16.957	0.63990	0.17516	139	1.48	005.99	0
005	Ash (%)	2126	17.960	0.06000	16.957	0.63990	0.17516	139	1.57	005.05	0
005	Ash (%)	0921	17.970	0.00000	16.957	0.63990	0.17516	139	1.58	005.05	0
005	Ash (%)	0941	18.005	0.19000	16.957	0.63990	0.17516	139	1.64	005.05	0
005	Ash (%)	0852	18.030	0.06000	16.957	0.63990	0.17516	139	1.68	005.00	0
005	Ash (%)	0685	18.100	0.02000	16.957	0.63990	0.17516	139	1.79	005.05	0
005	Ash (%)	2144	18.165	0.01000	16.957	0.63990	0.17516	139	1.89	005.99	0
005	Ash (%)	0652	18.250	0.10000	16.957	0.63990	0.17516	139	2.02	005.99	0
005	Ash (%)	0889	18.280	0.36000	16.957	0.63990	0.17516	139	2.07	005.05	0
005	Ash (%)	0910	17.750	1.1000	16.957	0.63990	0.17516	139	1.24	005.99	1
005	Ash (%)	2129	19.765	1.1500	16.957	0.63990	0.17516	139	4.39	005.05	1
005	Ash (%)	0889	25.675	0.61000	16.957	0.63990	0.17516	139	13.62	005.11	2
006	Total Sugars (%)	0939	2.4400	0.10000	3.5107	1.5456	0.31833	6	-0.69	006.99	0
006	Total Sugars (%)	0921	2.6400	0.00000	3.5107	1.5456	0.31833	6	-0.56	006.00	0
006	Total Sugars (%)	0910	2.6450	0.97000	3.5107	1.5456	0.31833	6	-0.56	006.00	0
006	Total Sugars (%)	0956	3.0000	0.20000	3.5107	1.5456	0.31833	6	-0.33	006.99	0
006	Total Sugars (%)	0407	4.5100	0.36000	3.5107	1.5456	0.31833	6	0.65	006.00	0
006	Total Sugars (%)	0187	17.670	0.28000	3.5107	1.5456	0.31833	6	9.16	006.00	0
008	Fiber, Acid Detergent (%)	0038	18.355	0.03000	22.423	1.5445	0.32888	55	-2.63	008.02	0
008	Fiber, Acid Detergent (%)	0956	19.400	1.0000	22.423	1.5445	0.32888	55	-1.96	008.08	0
008	Fiber, Acid Detergent (%)	0910	19.905	0.73000	22.423	1.5445	0.32888	55	-1.63	008.08	0
008	Fiber, Acid Detergent (%)	0939	20.090	0.44000	22.423	1.5445	0.32888	55	-1.51	008.08	0
008	Fiber, Acid Detergent (%)	0004	20.440	0.14000	22.423	1.5445	0.32888	55	-1.28	008.08	0
008	Fiber, Acid Detergent (%)	0590	20.555	0.29000	22.423	1.5445	0.32888	55	-1.21	008.02	0
008	Fiber, Acid Detergent (%)	2129	20.654	0.23200	22.423	1.5445	0.32888	55	-1.15	008.08	0
008	Fiber, Acid Detergent (%)	0510	20.750	0.30000	22.423	1.5445	0.32888	55	-1.08	008.08	0
008	Fiber, Acid Detergent (%)	0413	20.950	0.50000	22.423	1.5445	0.32888	55	-0.95	008.08	0
008	Fiber, Acid Detergent (%)	0878	20.980	0.18000	22.423	1.5445	0.32888	55	-0.93	008.08	0
008	Fiber, Acid Detergent (%)	0049	21.000	0.42000	22.423	1.5445	0.32888	55	-0.92	008.08	0
008	Fiber, Acid Detergent (%)	2190	21.170	0.54000	22.423	1.5445	0.32888	55	-0.81	008.08	0
008	Fiber, Acid Detergent (%)	0035	21.395	0.01000	22.423	1.5445	0.32888	55	-0.67	008.08	0
008	Fiber, Acid Detergent (%)	0889	21.490	0.02000	22.423	1.5445	0.32888	55	-0.60	008.99	0
008	Fiber, Acid Detergent (%)	2196	21.540	0.00000	22.423	1.5445	0.32888	55	-0.57	008.08	0
008	Fiber, Acid Detergent (%)	0066	21.585	0.24600	22.423	1.5445	0.32888	55	-0.54	008.08	0
008	Fiber, Acid Detergent (%)	0676	21.710	0.06000	22.423	1.5445	0.32888	55	-0.46	008.02	0
008	Fiber, Acid Detergent (%)	0889	21.715	0.05000	22.423	1.5445	0.32888	55	-0.46	008.08	0
008	Fiber, Acid Detergent (%)	0354	21.725	0.69000	22.423	1.5445	0.32888	55	-0.45	008.08	0
008	Fiber, Acid Detergent (%)	0581	21.825	0.13000	22.423	1.5445	0.32888	55	-0.39	008.08	0
008	Fiber, Acid Detergent (%)	0968	21.853	0.12900	22.423	1.5445	0.32888	55	-0.37	008.08	0
008	Fiber, Acid Detergent (%)	0610	21.950	0.30000	22.423	1.5445	0.32888	55	-0.31	008.99	0
008	Fiber, Acid Detergent (%)	0358	22.045	0.83000	22.423	1.5445	0.32888	55	-0.24	008.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
008	Fiber, Acid Detergent (%)	0357	22.050	0.10000	22.423	1.5445	0.32888	55	-0.24	008.08	0
008	Fiber, Acid Detergent (%)	2012	22.055	0.09000	22.423	1.5445	0.32888	55	-0.24	008.08	0
008	Fiber, Acid Detergent (%)	0948	22.060	0.54000	22.423	1.5445	0.32888	55	-0.23	008.02	0
008	Fiber, Acid Detergent (%)	0407	22.066	0.06600	22.423	1.5445	0.32888	55	-0.23	008.08	0
008	Fiber, Acid Detergent (%)	2192	22.160	0.52000	22.423	1.5445	0.32888	55	-0.17	008.02	0
008	Fiber, Acid Detergent (%)	0870	22.327	0.89480	22.423	1.5445	0.32888	55	-0.06	008.08	0
008	Fiber, Acid Detergent (%)	0941	22.500	0.16000	22.423	1.5445	0.32888	55	0.05	008.99	0
008	Fiber, Acid Detergent (%)	2193	22.715	0.51000	22.423	1.5445	0.32888	55	0.19	008.08	0
008	Fiber, Acid Detergent (%)	0504	22.875	0.07000	22.423	1.5445	0.32888	55	0.29	008.02	0
008	Fiber, Acid Detergent (%)	0148	22.880	0.74000	22.423	1.5445	0.32888	55	0.30	008.08	0
008	Fiber, Acid Detergent (%)	0353	22.940	0.08000	22.423	1.5445	0.32888	55	0.33	008.02	0
008	Fiber, Acid Detergent (%)	0619	23.040	0.00000	22.423	1.5445	0.32888	55	0.40	008.08	0
008	Fiber, Acid Detergent (%)	0226	23.050	0.50000	22.423	1.5445	0.32888	55	0.41	008.02	0
008	Fiber, Acid Detergent (%)	0278	23.150	0.70000	22.423	1.5445	0.32888	55	0.47	008.08	0
008	Fiber, Acid Detergent (%)	0098	23.230	0.12000	22.423	1.5445	0.32888	55	0.52	008.02	0
008	Fiber, Acid Detergent (%)	0148	23.240	0.36000	22.423	1.5445	0.32888	55	0.53	008.02	0
008	Fiber, Acid Detergent (%)	0265	23.400	0.20000	22.423	1.5445	0.32888	55	0.63	008.05	0
008	Fiber, Acid Detergent (%)	0918	23.406	0.68800	22.423	1.5445	0.32888	55	0.64	008.08	0
008	Fiber, Acid Detergent (%)	0001	23.525	0.15000	22.423	1.5445	0.32888	55	0.71	008.08	0
008	Fiber, Acid Detergent (%)	0037	23.600	0.08000	22.423	1.5445	0.32888	55	0.76	008.08	0
008	Fiber, Acid Detergent (%)	0646	23.620	0.62000	22.423	1.5445	0.32888	55	0.78	008.08	0
008	Fiber, Acid Detergent (%)	0345	23.650	0.30000	22.423	1.5445	0.32888	55	0.79	008.02	0
008	Fiber, Acid Detergent (%)	0083	23.865	0.37000	22.423	1.5445	0.32888	55	0.93	008.08	0
008	Fiber, Acid Detergent (%)	0045	23.900	0.60000	22.423	1.5445	0.32888	55	0.96	008.08	0
008	Fiber, Acid Detergent (%)	0187	23.915	0.15000	22.423	1.5445	0.32888	55	0.97	008.02	0
008	Fiber, Acid Detergent (%)	0693	24.432	0.22000	22.423	1.5445	0.32888	55	1.30	008.08	0
008	Fiber, Acid Detergent (%)	0190	24.505	0.21000	22.423	1.5445	0.32888	55	1.35	008.08	0
008	Fiber, Acid Detergent (%)	0309	24.517	0.68260	22.423	1.5445	0.32888	55	1.36	008.02	0
008	Fiber, Acid Detergent (%)	0689	24.600	0.00000	22.423	1.5445	0.32888	55	1.41	008.02	0
008	Fiber, Acid Detergent (%)	0294	25.250	0.50000	22.423	1.5445	0.32888	55	1.83	008.08	0
008	Fiber, Acid Detergent (%)	0675	26.840	0.60000	22.423	1.5445	0.32888	55	2.86	008.02	0
008	Fiber, Acid Detergent (%)	2066	27.500	0.00000	22.423	1.5445	0.32888	55	3.29	008.02	0
008	Fiber, Acid Detergent (%)	0861	22.180	2.5400	22.423	1.5445	0.32888	55	-0.16	008.08	1
008	Fiber, Acid Detergent (%)	0042	22.720	2.0400	22.423	1.5445	0.32888	55	0.19	008.08	1
009	Fiber, Neutral Detergent (%)	2012	32.475	0.27000	35.695	2.0018	0.41549	52	-1.61	009.09	0
009	Fiber, Neutral Detergent (%)	0939	32.765	1.0700	35.695	2.0018	0.41549	52	-1.46	009.09	0
009	Fiber, Neutral Detergent (%)	0918	32.774	0.52120	35.695	2.0018	0.41549	52	-1.46	009.09	0
009	Fiber, Neutral Detergent (%)	0878	32.825	0.21000	35.695	2.0018	0.41549	52	-1.43	009.09	0
009	Fiber, Neutral Detergent (%)	2022	33.165	0.27000	35.695	2.0018	0.41549	52	-1.26	009.09	0
009	Fiber, Neutral Detergent (%)	2193	33.170	0.14000	35.695	2.0018	0.41549	52	-1.26	009.09	0
009	Fiber, Neutral Detergent (%)	0148	33.245	0.29000	35.695	2.0018	0.41549	52	-1.22	009.09	0
009	Fiber, Neutral Detergent (%)	0407	33.366	0.13270	35.695	2.0018	0.41549	52	-1.16	009.09	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
009	Fiber, Neutral Detergent (%)	0581	33.395	0.19000	35.695	2.0018	0.41549	52	-1.15	009.09	0
009	Fiber, Neutral Detergent (%)	0861	33.625	1.1100	35.695	2.0018	0.41549	52	-1.03	009.09	0
009	Fiber, Neutral Detergent (%)	0294	33.800	0.00000	35.695	2.0018	0.41549	52	-0.95	009.09	0
009	Fiber, Neutral Detergent (%)	0598	34.185	0.43000	35.695	2.0018	0.41549	52	-0.75	009.09	0
009	Fiber, Neutral Detergent (%)	0889	34.210	0.06000	35.695	2.0018	0.41549	52	-0.74	009.99	0
009	Fiber, Neutral Detergent (%)	0357	34.300	0.20000	35.695	2.0018	0.41549	52	-0.70	009.09	0
009	Fiber, Neutral Detergent (%)	0956	34.300	0.00000	35.695	2.0018	0.41549	52	-0.70	009.09	0
009	Fiber, Neutral Detergent (%)	0510	34.400	1.0000	35.695	2.0018	0.41549	52	-0.65	009.09	0
009	Fiber, Neutral Detergent (%)	0413	34.500	0.20000	35.695	2.0018	0.41549	52	-0.60	009.09	0
009	Fiber, Neutral Detergent (%)	2196	34.720	0.00000	35.695	2.0018	0.41549	52	-0.49	009.09	0
009	Fiber, Neutral Detergent (%)	0870	35.347	0.97430	35.695	2.0018	0.41549	52	-0.17	009.09	0
009	Fiber, Neutral Detergent (%)	2190	35.375	0.87000	35.695	2.0018	0.41549	52	-0.16	009.09	0
009	Fiber, Neutral Detergent (%)	0910	35.400	1.0000	35.695	2.0018	0.41549	52	-0.15	009.07	0
009	Fiber, Neutral Detergent (%)	0646	35.420	0.14000	35.695	2.0018	0.41549	52	-0.14	009.09	0
009	Fiber, Neutral Detergent (%)	0098	35.500	0.60000	35.695	2.0018	0.41549	52	-0.10	009.07	0
009	Fiber, Neutral Detergent (%)	0190	35.550	0.26000	35.695	2.0018	0.41549	52	-0.07	009.09	0
009	Fiber, Neutral Detergent (%)	2129	36.037	0.64220	35.695	2.0018	0.41549	52	0.17	009.09	0
009	Fiber, Neutral Detergent (%)	0226	36.050	0.10000	35.695	2.0018	0.41549	52	0.18	009.07	0
009	Fiber, Neutral Detergent (%)	0083	36.080	1.0600	35.695	2.0018	0.41549	52	0.19	009.09	0
009	Fiber, Neutral Detergent (%)	0889	36.160	0.34000	35.695	2.0018	0.41549	52	0.23	009.09	0
009	Fiber, Neutral Detergent (%)	2192	36.175	0.47000	35.695	2.0018	0.41549	52	0.24	009.07	0
009	Fiber, Neutral Detergent (%)	0045	36.200	0.20000	35.695	2.0018	0.41549	52	0.25	009.09	0
009	Fiber, Neutral Detergent (%)	0037	36.390	0.22000	35.695	2.0018	0.41549	52	0.35	009.09	0
009	Fiber, Neutral Detergent (%)	0187	36.495	0.09000	35.695	2.0018	0.41549	52	0.40	009.07	0
009	Fiber, Neutral Detergent (%)	0353	36.505	0.65000	35.695	2.0018	0.41549	52	0.40	009.07	0
009	Fiber, Neutral Detergent (%)	0693	36.593	0.23400	35.695	2.0018	0.41549	52	0.45	009.09	0
009	Fiber, Neutral Detergent (%)	0345	36.620	0.40000	35.695	2.0018	0.41549	52	0.46	009.07	0
009	Fiber, Neutral Detergent (%)	2066	36.655	0.67000	35.695	2.0018	0.41549	52	0.48	009.07	0
009	Fiber, Neutral Detergent (%)	0066	36.697	0.22600	35.695	2.0018	0.41549	52	0.50	009.09	0
009	Fiber, Neutral Detergent (%)	0968	36.851	0.12400	35.695	2.0018	0.41549	52	0.58	009.09	0
009	Fiber, Neutral Detergent (%)	0590	36.900	1.3000	35.695	2.0018	0.41549	52	0.60	009.07	0
009	Fiber, Neutral Detergent (%)	0676	36.915	0.55000	35.695	2.0018	0.41549	52	0.61	009.07	0
009	Fiber, Neutral Detergent (%)	0948	36.940	0.40000	35.695	2.0018	0.41549	52	0.62	009.09	0
009	Fiber, Neutral Detergent (%)	0265	36.950	0.10000	35.695	2.0018	0.41549	52	0.63	009.09	0
009	Fiber, Neutral Detergent (%)	0001	37.295	0.71000	35.695	2.0018	0.41549	52	0.80	009.09	0
009	Fiber, Neutral Detergent (%)	0504	37.410	0.42000	35.695	2.0018	0.41549	52	0.86	009.04	0
009	Fiber, Neutral Detergent (%)	0354	37.545	0.29000	35.695	2.0018	0.41549	52	0.92	009.09	0
009	Fiber, Neutral Detergent (%)	0278	37.600	0.00000	35.695	2.0018	0.41549	52	0.95	009.09	0
009	Fiber, Neutral Detergent (%)	0610	38.050	0.10000	35.695	2.0018	0.41549	52	1.18	009.07	0
009	Fiber, Neutral Detergent (%)	0309	38.201	0.31110	35.695	2.0018	0.41549	52	1.25	009.07	0
009	Fiber, Neutral Detergent (%)	0941	38.865	0.61000	35.695	2.0018	0.41549	52	1.58	009.99	0
009	Fiber, Neutral Detergent (%)	0619	39.350	0.10000	35.695	2.0018	0.41549	52	1.83	009.09	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
009	Fiber, Neutral Detergent (%)	0689	39.650	0.50000	35.695	2.0018	0.41549	52	1.98	009.07	0
009	Fiber, Neutral Detergent (%)	0675	41.585	0.85000	35.695	2.0018	0.41549	52	2.94	009.07	0
009	Fiber, Neutral Detergent (%)	0049	32.715	2.7900	35.695	2.0018	0.41549	52	-1.49	009.09	1
009	Fiber, Neutral Detergent (%)	0358	34.920	3.3200	35.695	2.0018	0.41549	52	-0.39	009.09	1
010	Moisture (%)	0889	7.0300	0.02000	9.2446	0.80051	0.12397	23	-2.77	010.11	0
010	Moisture (%)	0852	8.2250	0.37000	9.2446	0.80051	0.12397	23	-1.27	010.11	0
010	Moisture (%)	2129	8.3052	0.19030	9.2446	0.80051	0.12397	23	-1.17	010.99	0
010	Moisture (%)	0843	8.5850	0.13000	9.2446	0.80051	0.12397	23	-0.82	010.03	0
010	Moisture (%)	0164	8.6350	0.11000	9.2446	0.80051	0.12397	23	-0.76	010.03	0
010	Moisture (%)	2150	8.7050	0.07000	9.2446	0.80051	0.12397	23	-0.67	010.99	0
010	Moisture (%)	2144	8.8500	0.04000	9.2446	0.80051	0.12397	23	-0.49	010.99	0
010	Moisture (%)	0853	8.9000	0.20000	9.2446	0.80051	0.12397	23	-0.43	010.99	0
010	Moisture (%)	0722	8.9467	0.00590	9.2446	0.80051	0.12397	23	-0.37	010.99	0
010	Moisture (%)	0884	9.0000	0.20000	9.2446	0.80051	0.12397	23	-0.31	010.99	0
010	Moisture (%)	0948	9.0100	0.02000	9.2446	0.80051	0.12397	23	-0.29	010.99	0
010	Moisture (%)	0652	9.2000	0.00000	9.2446	0.80051	0.12397	23	-0.06	010.99	0
010	Moisture (%)	2105	9.2750	0.05000	9.2446	0.80051	0.12397	23	0.04	010.11	0
010	Moisture (%)	0917	9.3000	0.20000	9.2446	0.80051	0.12397	23	0.07	010.99	0
010	Moisture (%)	0553	9.4550	0.09000	9.2446	0.80051	0.12397	23	0.26	010.11	0
010	Moisture (%)	0939	9.5200	0.48000	9.2446	0.80051	0.12397	23	0.34	010.99	0
010	Moisture (%)	0190	9.6000	0.04000	9.2446	0.80051	0.12397	23	0.44	010.99	0
010	Moisture (%)	0921	9.6600	0.02000	9.2446	0.80051	0.12397	23	0.52	010.99	0
010	Moisture (%)	2089	9.8950	0.53000	9.2446	0.80051	0.12397	23	0.81	010.99	0
010	Moisture (%)	0152	10.240	0.02000	9.2446	0.80051	0.12397	23	1.24	010.11	0
010	Moisture (%)	0952	10.385	0.03000	9.2446	0.80051	0.12397	23	1.42	010.99	0
010	Moisture (%)	2128	11.060	0.03500	9.2446	0.80051	0.12397	23	2.27	010.11	0
010	Moisture (%)	0405	11.260	0.00000	9.2446	0.80051	0.12397	23	2.52	010.99	0
011	Loss on Drying (%)	0643	8.1100	0.16000	10.102	0.45797	0.17092	76	-4.35	011.01	0
011	Loss on Drying (%)	0618	8.4350	0.05000	10.102	0.45797	0.17092	76	-3.64	011.01	0
011	Loss on Drying (%)	0175	8.9500	0.30000	10.102	0.45797	0.17092	76	-2.51	011.01	0
011	Loss on Drying (%)	2199	8.9500	0.10000	10.102	0.45797	0.17092	76	-2.51	011.01	0
011	Loss on Drying (%)	0853	9.1000	0.00000	10.102	0.45797	0.17092	76	-2.19	011.01	0
011	Loss on Drying (%)	2031	9.2691	0.18810	10.102	0.45797	0.17092	76	-1.82	011.01	0
011	Loss on Drying (%)	0874	9.3000	0.02000	10.102	0.45797	0.17092	76	-1.75	011.01	0
011	Loss on Drying (%)	2073	9.3050	0.29000	10.102	0.45797	0.17092	76	-1.74	011.01	0
011	Loss on Drying (%)	2200	9.4500	0.70000	10.102	0.45797	0.17092	76	-1.42	011.01	0
011	Loss on Drying (%)	0510	9.5000	0.00000	10.102	0.45797	0.17092	76	-1.31	011.01	0
011	Loss on Drying (%)	0859	9.5135	0.07700	10.102	0.45797	0.17092	76	-1.28	011.01	0
011	Loss on Drying (%)	0298	9.5600	0.02000	10.102	0.45797	0.17092	76	-1.18	011.01	0
011	Loss on Drying (%)	0953	9.5650	0.79000	10.102	0.45797	0.17092	76	-1.17	011.01	0
011	Loss on Drying (%)	0536	9.5900	0.02000	10.102	0.45797	0.17092	76	-1.12	011.01	0
011	Loss on Drying (%)	0880	9.6480	1.4780	10.102	0.45797	0.17092	76	-0.99	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
011	Loss on Drying (%)	0942	9.6500	0.22000	10.102	0.45797	0.17092	76	-0.99	011.02	0
011	Loss on Drying (%)	0860	9.7600	0.00000	10.102	0.45797	0.17092	76	-0.75	011.01	0
011	Loss on Drying (%)	2168	9.7850	0.09000	10.102	0.45797	0.17092	76	-0.69	011.01	0
011	Loss on Drying (%)	0968	9.8110	0.00400	10.102	0.45797	0.17092	76	-0.63	011.01	0
011	Loss on Drying (%)	0354	9.8550	0.01000	10.102	0.45797	0.17092	76	-0.54	011.01	0
011	Loss on Drying (%)	0539	9.9200	0.42000	10.102	0.45797	0.17092	76	-0.40	011.01	0
011	Loss on Drying (%)	0108	9.9300	0.66000	10.102	0.45797	0.17092	76	-0.37	011.01	0
011	Loss on Drying (%)	0100	9.9500	0.00000	10.102	0.45797	0.17092	76	-0.33	011.01	0
011	Loss on Drying (%)	0265	9.9500	0.10000	10.102	0.45797	0.17092	76	-0.33	011.01	0
011	Loss on Drying (%)	0294	9.9500	0.10000	10.102	0.45797	0.17092	76	-0.33	011.01	0
011	Loss on Drying (%)	0164	9.9800	0.04000	10.102	0.45797	0.17092	76	-0.27	011.01	0
011	Loss on Drying (%)	2075	9.9900	0.08000	10.102	0.45797	0.17092	76	-0.24	011.02	0
011	Loss on Drying (%)	0622	10.038	0.00400	10.102	0.45797	0.17092	76	-0.14	011.01	0
011	Loss on Drying (%)	0226	10.040	0.02000	10.102	0.45797	0.17092	76	-0.13	011.01	0
011	Loss on Drying (%)	0857	10.050	0.10000	10.102	0.45797	0.17092	76	-0.11	011.99	0
011	Loss on Drying (%)	0233	10.055	0.23000	10.102	0.45797	0.17092	76	-0.10	011.01	0
011	Loss on Drying (%)	0653	10.055	0.05000	10.102	0.45797	0.17092	76	-0.10	011.01	0
011	Loss on Drying (%)	0870	10.085	0.06600	10.102	0.45797	0.17092	76	-0.04	011.01	0
011	Loss on Drying (%)	0596	10.100	0.04000	10.102	0.45797	0.17092	76	0.00	011.01	0
011	Loss on Drying (%)	0685	10.100	0.18000	10.102	0.45797	0.17092	76	0.00	011.01	0
011	Loss on Drying (%)	0208	10.110	0.58000	10.102	0.45797	0.17092	76	0.02	011.01	0
011	Loss on Drying (%)	0242	10.120	0.14000	10.102	0.45797	0.17092	76	0.04	011.01	0
011	Loss on Drying (%)	0553	10.135	0.17000	10.102	0.45797	0.17092	76	0.07	011.01	0
011	Loss on Drying (%)	0407	10.138	0.00550	10.102	0.45797	0.17092	76	0.08	011.01	0
011	Loss on Drying (%)	0202	10.155	0.13000	10.102	0.45797	0.17092	76	0.12	011.01	0
011	Loss on Drying (%)	0541	10.170	0.30000	10.102	0.45797	0.17092	76	0.15	011.01	0
011	Loss on Drying (%)	0229	10.197	0.02900	10.102	0.45797	0.17092	76	0.21	011.01	0
011	Loss on Drying (%)	0520	10.200	0.00000	10.102	0.45797	0.17092	76	0.21	011.01	0
011	Loss on Drying (%)	2196	10.200	0.00000	10.102	0.45797	0.17092	76	0.21	011.01	0
011	Loss on Drying (%)	2097	10.204	0.02560	10.102	0.45797	0.17092	76	0.22	011.01	0
011	Loss on Drying (%)	2022	10.205	0.01000	10.102	0.45797	0.17092	76	0.23	011.01	0
011	Loss on Drying (%)	0511	10.230	0.02000	10.102	0.45797	0.17092	76	0.28	011.01	0
011	Loss on Drying (%)	0425	10.235	0.07000	10.102	0.45797	0.17092	76	0.29	011.02	0
011	Loss on Drying (%)	0171	10.240	0.02000	10.102	0.45797	0.17092	76	0.30	011.01	0
011	Loss on Drying (%)	0646	10.265	0.15000	10.102	0.45797	0.17092	76	0.36	011.01	0
011	Loss on Drying (%)	0358	10.295	0.23000	10.102	0.45797	0.17092	76	0.42	011.01	0
011	Loss on Drying (%)	0598	10.295	0.27000	10.102	0.45797	0.17092	76	0.42	011.01	0
011	Loss on Drying (%)	0033	10.300	0.00000	10.102	0.45797	0.17092	76	0.43	011.01	0
011	Loss on Drying (%)	0675	10.325	0.09000	10.102	0.45797	0.17092	76	0.49	011.01	0
011	Loss on Drying (%)	0144	10.380	0.12000	10.102	0.45797	0.17092	76	0.61	011.01	0
011	Loss on Drying (%)	0682	10.380	0.00000	10.102	0.45797	0.17092	76	0.61	011.01	0
011	Loss on Drying (%)	0589	10.390	0.14000	10.102	0.45797	0.17092	76	0.63	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
011	Loss on Drying (%)	0574	10.425	0.07000	10.102	0.45797	0.17092	76	0.71	011.01	0
011	Loss on Drying (%)	0309	10.440	0.08000	10.102	0.45797	0.17092	76	0.74	011.01	0
011	Loss on Drying (%)	2006	10.450	0.12000	10.102	0.45797	0.17092	76	0.76	011.01	0
011	Loss on Drying (%)	2188	10.465	0.07000	10.102	0.45797	0.17092	76	0.79	011.01	0
011	Loss on Drying (%)	0098	10.475	0.33000	10.102	0.45797	0.17092	76	0.82	011.01	0
011	Loss on Drying (%)	0861	10.500	0.10000	10.102	0.45797	0.17092	76	0.87	011.01	0
011	Loss on Drying (%)	2193	10.505	0.05000	10.102	0.45797	0.17092	76	0.88	011.01	0
011	Loss on Drying (%)	0843	10.515	0.07000	10.102	0.45797	0.17092	76	0.90	011.02	0
011	Loss on Drying (%)	0881	10.538	0.06800	10.102	0.45797	0.17092	76	0.95	011.01	0
011	Loss on Drying (%)	0651	10.547	0.01200	10.102	0.45797	0.17092	76	0.97	011.01	0
011	Loss on Drying (%)	0066	10.632	0.06800	10.102	0.45797	0.17092	76	1.16	011.01	0
011	Loss on Drying (%)	0004	10.640	0.10000	10.102	0.45797	0.17092	76	1.18	011.01	0
011	Loss on Drying (%)	0650	10.660	0.24000	10.102	0.45797	0.17092	76	1.22	011.01	0
011	Loss on Drying (%)	0722	10.686	0.04500	10.102	0.45797	0.17092	76	1.28	011.01	0
011	Loss on Drying (%)	0047	10.750	0.10000	10.102	0.45797	0.17092	76	1.42	011.01	0
011	Loss on Drying (%)	0123	10.765	0.01000	10.102	0.45797	0.17092	76	1.45	011.01	0
011	Loss on Drying (%)	0205	10.820	0.10000	10.102	0.45797	0.17092	76	1.57	011.01	0
011	Loss on Drying (%)	2181	11.105	2.0300	10.102	0.45797	0.17092	76	2.19	011.01	0
011	Loss on Drying (%)	0852	11.280	0.22000	10.102	0.45797	0.17092	76	2.57	011.99	0
011	Loss on Drying (%)	0674	5.7100	9.0000	10.102	0.45797	0.17092	76	-9.59	011.01	2
012	Starch (%)	0510	5.6500	0.10000	9.1592	0.99944	0.24236	37	-3.51	012.04	0
012	Starch (%)	0852	7.0550	0.39000	9.1592	0.99944	0.24236	37	-2.11	012.11	0
012	Starch (%)	0918	7.1640	0.69600	9.1592	0.99944	0.24236	37	-2.00	012.01	0
012	Starch (%)	0676	7.8350	0.13000	9.1592	0.99944	0.24236	37	-1.32	012.01	0
012	Starch (%)	0160	8.0250	0.51000	9.1592	0.99944	0.24236	37	-1.13	012.04	0
012	Starch (%)	0098	8.1100	0.08000	9.1592	0.99944	0.24236	37	-1.05	012.03	0
012	Starch (%)	0164	8.2500	0.10000	9.1592	0.99944	0.24236	37	-0.91	012.01	0
012	Starch (%)	0956	8.2500	0.10000	9.1592	0.99944	0.24236	37	-0.91	012.01	0
012	Starch (%)	0164	8.4000	0.00000	9.1592	0.99944	0.24236	37	-0.76	012.00	0
012	Starch (%)	0407	8.4526	0.35480	9.1592	0.99944	0.24236	37	-0.71	012.03	0
012	Starch (%)	0278	8.5500	0.30000	9.1592	0.99944	0.24236	37	-0.61	012.04	0
012	Starch (%)	0045	8.5700	0.14000	9.1592	0.99944	0.24236	37	-0.59	012.01	0
012	Starch (%)	0353	8.6950	0.09000	9.1592	0.99944	0.24236	37	-0.46	012.04	0
012	Starch (%)	0889	9.0050	0.23000	9.1592	0.99944	0.24236	37	-0.15	012.11	0
012	Starch (%)	0939	9.0450	0.05000	9.1592	0.99944	0.24236	37	-0.11	012.99	0
012	Starch (%)	0948	9.0550	0.53000	9.1592	0.99944	0.24236	37	-0.10	012.04	0
012	Starch (%)	0910	9.1000	1.0000	9.1592	0.99944	0.24236	37	-0.06	012.00	0
012	Starch (%)	0870	9.1252	0.78960	9.1592	0.99944	0.24236	37	-0.03	012.01	0
012	Starch (%)	0693	9.1430	0.05600	9.1592	0.99944	0.24236	37	-0.02	012.01	0
012	Starch (%)	0148	9.2650	0.19000	9.1592	0.99944	0.24236	37	0.11	012.01	0
012	Starch (%)	2193	9.2800	0.10000	9.1592	0.99944	0.24236	37	0.12	012.00	0
012	Starch (%)	2195	9.3000	0.10000	9.1592	0.99944	0.24236	37	0.14	012.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
012	Starch (%)	2181	9.5150	0.11000	9.1592	0.99944	0.24236	37	0.36	012.00	0
012	Starch (%)	0689	9.6500	0.30000	9.1592	0.99944	0.24236	37	0.49	012.00	0
012	Starch (%)	2196	9.7000	0.00000	9.1592	0.99944	0.24236	37	0.54	012.00	0
012	Starch (%)	2188	9.7350	0.03000	9.1592	0.99944	0.24236	37	0.58	012.00	0
012	Starch (%)	0619	9.7650	0.21000	9.1592	0.99944	0.24236	37	0.61	012.00	0
012	Starch (%)	2006	9.9050	0.07000	9.1592	0.99944	0.24236	37	0.75	012.00	0
012	Starch (%)	0610	9.9500	0.10000	9.1592	0.99944	0.24236	37	0.79	012.00	0
012	Starch (%)	0354	9.9900	0.02000	9.1592	0.99944	0.24236	37	0.83	012.00	0
012	Starch (%)	0941	10.000	0.40000	9.1592	0.99944	0.24236	37	0.84	012.00	0
012	Starch (%)	2089	10.005	0.35000	9.1592	0.99944	0.24236	37	0.85	012.01	0
012	Starch (%)	0878	10.035	0.15000	9.1592	0.99944	0.24236	37	0.88	012.00	0
012	Starch (%)	2009	10.231	0.02000	9.1592	0.99944	0.24236	37	1.07	012.00	0
012	Starch (%)	2192	10.740	0.52000	9.1592	0.99944	0.24236	37	1.58	012.00	0
012	Starch (%)	0265	11.150	0.30000	9.1592	0.99944	0.24236	37	1.99	012.01	0
012	Starch (%)	2128	13.112	0.35100	9.1592	0.99944	0.24236	37	3.95	012.11	0
013	Fat, Acid Pretreat (%)	2128	2.0278	0.10000	3.2398	0.60026	0.12093	47	-2.02	013.12	0
013	Fat, Acid Pretreat (%)	2146	2.2000	0.20000	3.2398	0.60026	0.12093	47	-1.73	013.00	0
013	Fat, Acid Pretreat (%)	0504	2.3650	0.33000	3.2398	0.60026	0.12093	47	-1.46	013.00	0
013	Fat, Acid Pretreat (%)	2145	2.4100	0.02000	3.2398	0.60026	0.12093	47	-1.38	013.00	0
013	Fat, Acid Pretreat (%)	2196	2.5000	0.00000	3.2398	0.60026	0.12093	47	-1.23	013.10	0
013	Fat, Acid Pretreat (%)	0036	2.5524	0.01010	3.2398	0.60026	0.12093	47	-1.15	013.10	0
013	Fat, Acid Pretreat (%)	0083	2.5950	0.11000	3.2398	0.60026	0.12093	47	-1.07	013.00	0
013	Fat, Acid Pretreat (%)	2006	2.6100	0.06000	3.2398	0.60026	0.12093	47	-1.05	013.10	0
013	Fat, Acid Pretreat (%)	0689	2.6500	0.10000	3.2398	0.60026	0.12093	47	-0.98	013.00	0
013	Fat, Acid Pretreat (%)	0618	2.7435	0.40780	3.2398	0.60026	0.12093	47	-0.83	013.08	0
013	Fat, Acid Pretreat (%)	0643	2.8150	0.17000	3.2398	0.60026	0.12093	47	-0.71	013.02	0
013	Fat, Acid Pretreat (%)	0407	2.8206	0.13910	3.2398	0.60026	0.12093	47	-0.70	013.13	0
013	Fat, Acid Pretreat (%)	0638	2.8500	0.10000	3.2398	0.60026	0.12093	47	-0.65	013.10	0
013	Fat, Acid Pretreat (%)	0910	2.9550	0.27000	3.2398	0.60026	0.12093	47	-0.47	013.00	0
013	Fat, Acid Pretreat (%)	0948	2.9600	0.04000	3.2398	0.60026	0.12093	47	-0.47	013.00	0
013	Fat, Acid Pretreat (%)	0652	3.0000	0.00000	3.2398	0.60026	0.12093	47	-0.40	013.00	0
013	Fat, Acid Pretreat (%)	0884	3.0000	0.00000	3.2398	0.60026	0.12093	47	-0.40	013.00	0
013	Fat, Acid Pretreat (%)	0874	3.0000	0.20000	3.2398	0.60026	0.12093	47	-0.40	013.02	0
013	Fat, Acid Pretreat (%)	0610	3.0350	0.27000	3.2398	0.60026	0.12093	47	-0.34	013.10	0
013	Fat, Acid Pretreat (%)	0853	3.0450	0.11000	3.2398	0.60026	0.12093	47	-0.32	013.02	0
013	Fat, Acid Pretreat (%)	0098	3.0900	0.10000	3.2398	0.60026	0.12093	47	-0.25	013.02	0
013	Fat, Acid Pretreat (%)	2073	3.1900	0.02000	3.2398	0.60026	0.12093	47	-0.08	013.00	0
013	Fat, Acid Pretreat (%)	0309	3.2419	0.04490	3.2398	0.60026	0.12093	47	0.00	013.00	0
013	Fat, Acid Pretreat (%)	0148	3.2800	0.16000	3.2398	0.60026	0.12093	47	0.07	013.02	0
013	Fat, Acid Pretreat (%)	0541	3.2950	0.07000	3.2398	0.60026	0.12093	47	0.09	013.00	0
013	Fat, Acid Pretreat (%)	0208	3.3000	0.46000	3.2398	0.60026	0.12093	47	0.10	013.02	0
013	Fat, Acid Pretreat (%)	0229	3.3100	0.02000	3.2398	0.60026	0.12093	47	0.12	013.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
013	Fat, Acid Pretreat (%)	0265	3.3300	0.08000	3.2398	0.60026	0.12093	47	0.15	013.13	0
013	Fat, Acid Pretreat (%)	0955	3.3750	0.07000	3.2398	0.60026	0.12093	47	0.23	013.13	0
013	Fat, Acid Pretreat (%)	0553	3.4800	0.10000	3.2398	0.60026	0.12093	47	0.40	013.02	0
013	Fat, Acid Pretreat (%)	0354	3.5100	0.02000	3.2398	0.60026	0.12093	47	0.45	013.02	0
013	Fat, Acid Pretreat (%)	0682	3.5300	0.00000	3.2398	0.60026	0.12093	47	0.48	013.02	0
013	Fat, Acid Pretreat (%)	0676	3.5350	0.07000	3.2398	0.60026	0.12093	47	0.49	013.02	0
013	Fat, Acid Pretreat (%)	0870	3.5495	0.19190	3.2398	0.60026	0.12093	47	0.52	013.02	0
013	Fat, Acid Pretreat (%)	0100	3.6250	0.19000	3.2398	0.60026	0.12093	47	0.64	013.02	0
013	Fat, Acid Pretreat (%)	0139	3.6950	0.01000	3.2398	0.60026	0.12093	47	0.76	013.00	0
013	Fat, Acid Pretreat (%)	0164	3.7300	0.04000	3.2398	0.60026	0.12093	47	0.82	013.02	0
013	Fat, Acid Pretreat (%)	0650	3.7750	0.07000	3.2398	0.60026	0.12093	47	0.89	013.02	0
013	Fat, Acid Pretreat (%)	2150	3.8350	0.29000	3.2398	0.60026	0.12093	47	0.99	013.00	0
013	Fat, Acid Pretreat (%)	0202	3.8450	0.25000	3.2398	0.60026	0.12093	47	1.01	013.00	0
013	Fat, Acid Pretreat (%)	0843	3.8500	0.04000	3.2398	0.60026	0.12093	47	1.02	013.13	0
013	Fat, Acid Pretreat (%)	0921	3.8600	0.12000	3.2398	0.60026	0.12093	47	1.03	013.02	0
013	Fat, Acid Pretreat (%)	0675	3.9800	0.28000	3.2398	0.60026	0.12093	47	1.23	013.02	0
013	Fat, Acid Pretreat (%)	0581	4.0550	0.07000	3.2398	0.60026	0.12093	47	1.36	013.13	0
013	Fat, Acid Pretreat (%)	0861	4.3550	0.15000	3.2398	0.60026	0.12093	47	1.86	013.02	0
013	Fat, Acid Pretreat (%)	0353	4.5900	0.12000	3.2398	0.60026	0.12093	47	2.25	013.10	0
013	Fat, Acid Pretreat (%)	0425	5.3850	0.01000	3.2398	0.60026	0.12093	47	3.57	013.02	0
013	Fat, Acid Pretreat (%)	0242	3.1000	0.68000	3.2398	0.60026	0.12093	47	-0.23	013.00	1
015	Aluminum (mg / kg (ppm))	0353	177.50	5.0000	274.58	34.835	10.894	13	-2.79	015.43	0
015	Aluminum (mg / kg (ppm))	0186	215.50	13.000	274.58	34.835	10.894	13	-1.70	015.52	0
015	Aluminum (mg / kg (ppm))	0049	234.98	10.640	274.58	34.835	10.894	13	-1.14	015.41	0
015	Aluminum (mg / kg (ppm))	0918	266.91	15.430	274.58	34.835	10.894	13	-0.22	015.43	0
015	Aluminum (mg / kg (ppm))	0510	271.50	1.0000	274.58	34.835	10.894	13	-0.09	015.43	0
015	Aluminum (mg / kg (ppm))	0345	281.00	6.0000	274.58	34.835	10.894	13	0.18	015.43	0
015	Aluminum (mg / kg (ppm))	0560	285.30	7.6000	274.58	34.835	10.894	13	0.31	015.42	0
015	Aluminum (mg / kg (ppm))	0169	285.50	11.000	274.58	34.835	10.894	13	0.31	015.43	0
015	Aluminum (mg / kg (ppm))	0171	288.73	11.950	274.58	34.835	10.894	13	0.41	015.41	0
015	Aluminum (mg / kg (ppm))	0218	290.00	20.000	274.58	34.835	10.894	13	0.44	015.43	0
015	Aluminum (mg / kg (ppm))	0164	295.00	10.000	274.58	34.835	10.894	13	0.59	015.41	0
015	Aluminum (mg / kg (ppm))	0520	304.50	23.000	274.58	34.835	10.894	13	0.86	015.41	0
015	Aluminum (mg / kg (ppm))	0553	321.50	7.0000	274.58	34.835	10.894	13	1.35	015.53	0
017	Boron (mg / kg (ppm))	0358	9.3350	0.07000	10.655	1.2814	0.42771	14	-1.03	017.41	0
017	Boron (mg / kg (ppm))	0560	9.5550	0.39000	10.655	1.2814	0.42771	14	-0.86	017.42	0
017	Boron (mg / kg (ppm))	0045	9.5800	0.10000	10.655	1.2814	0.42771	14	-0.84	017.42	0
017	Boron (mg / kg (ppm))	0229	9.7400	1.0600	10.655	1.2814	0.42771	14	-0.71	017.41	0
017	Boron (mg / kg (ppm))	0693	9.8780	0.49800	10.655	1.2814	0.42771	14	-0.61	017.42	0
017	Boron (mg / kg (ppm))	0345	9.9500	0.30000	10.655	1.2814	0.42771	14	-0.55	017.43	0
017	Boron (mg / kg (ppm))	2053	10.310	0.94000	10.655	1.2814	0.42771	14	-0.27	017.42	0
017	Boron (mg / kg (ppm))	0918	10.495	0.01000	10.655	1.2814	0.42771	14	-0.12	017.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
017	Boron (mg / kg (ppm))	0226	10.540	0.52000	10.655	1.2814	0.42771	14	-0.09	017.41	0
017	Boron (mg / kg (ppm))	0083	11.000	0.00000	10.655	1.2814	0.42771	14	0.27	017.43	0
017	Boron (mg / kg (ppm))	0353	11.500	1.0000	10.655	1.2814	0.42771	14	0.66	017.43	0
017	Boron (mg / kg (ppm))	0049	12.130	0.72000	10.655	1.2814	0.42771	14	1.15	017.41	0
017	Boron (mg / kg (ppm))	0510	13.000	0.00000	10.655	1.2814	0.42771	14	1.83	017.43	0
017	Boron (mg / kg (ppm))	0294	13.690	0.38000	10.655	1.2814	0.42771	14	2.37	017.42	0
017	Boron (mg / kg (ppm))	0265	8.7300	5.5400	10.655	1.2814	0.42771	14	-1.50	017.42	1
017	Boron (mg / kg (ppm))	0032	0.00000	0.00000	10.655	1.2814	0.42771	14		017.53	4
017	Boron (mg / kg (ppm))	0218	< 25		10.655	1.2814	0.42771	14		017.43	5
019	Calcium (%)	2199	1.7300	0.02000	3.4502	0.17245	0.08560	141	-9.97	019.08	0
019	Calcium (%)	2128	2.9390	0.33800	3.4502	0.17245	0.08560	141	-2.96	019.02	0
019	Calcium (%)	2055	2.9651	0.00120	3.4502	0.17245	0.08560	141	-2.81	019.53	0
019	Calcium (%)	2031	3.0290	0.02170	3.4502	0.17245	0.08560	141	-2.44	019.08	0
019	Calcium (%)	2200	3.0500	0.30000	3.4502	0.17245	0.08560	141	-2.32	019.08	0
019	Calcium (%)	2192	3.0950	0.07000	3.4502	0.17245	0.08560	141	-2.06	019.43	0
019	Calcium (%)	0098	3.1000	0.06000	3.4502	0.17245	0.08560	141	-2.03	019.41	0
019	Calcium (%)	0042	3.1350	0.09000	3.4502	0.17245	0.08560	141	-1.83	019.42	0
019	Calcium (%)	0144	3.1450	0.13000	3.4502	0.17245	0.08560	141	-1.77	019.41	0
019	Calcium (%)	0300	3.1650	0.19000	3.4502	0.17245	0.08560	141	-1.65	019.53	0
019	Calcium (%)	0511	3.1750	0.51000	3.4502	0.17245	0.08560	141	-1.60	019.41	0
019	Calcium (%)	2145	3.1760	0.03050	3.4502	0.17245	0.08560	141	-1.59	019.41	0
019	Calcium (%)	0175	3.1800	0.12000	3.4502	0.17245	0.08560	141	-1.57	019.31	0
019	Calcium (%)	0590	3.2250	0.01000	3.4502	0.17245	0.08560	141	-1.31	019.99	0
019	Calcium (%)	0345	3.2350	0.07000	3.4502	0.17245	0.08560	141	-1.25	019.43	0
019	Calcium (%)	0164	3.2450	0.05000	3.4502	0.17245	0.08560	141	-1.19	019.41	0
019	Calcium (%)	0038	3.2450	0.05000	3.4502	0.17245	0.08560	141	-1.19	019.43	0
019	Calcium (%)	0610	3.2650	0.07000	3.4502	0.17245	0.08560	141	-1.07	019.43	0
019	Calcium (%)	0036	3.2652	0.01270	3.4502	0.17245	0.08560	141	-1.07	019.43	0
019	Calcium (%)	0889	3.2669	0.00090	3.4502	0.17245	0.08560	141	-1.06	019.99	0
019	Calcium (%)	0011	3.2800	0.00800	3.4502	0.17245	0.08560	141	-0.99	019.41	0
019	Calcium (%)	0619	3.2800	0.06000	3.4502	0.17245	0.08560	141	-0.99	019.41	0
019	Calcium (%)	0536	3.2850	0.07000	3.4502	0.17245	0.08560	141	-0.96	019.31	0
019	Calcium (%)	0861	3.2850	0.19000	3.4502	0.17245	0.08560	141	-0.96	019.43	0
019	Calcium (%)	0596	3.2900	0.02000	3.4502	0.17245	0.08560	141	-0.93	019.31	0
019	Calcium (%)	0202	3.2960	0.07600	3.4502	0.17245	0.08560	141	-0.89	019.43	0
019	Calcium (%)	0405	3.3000	0.00000	3.4502	0.17245	0.08560	141	-0.87	019.41	0
019	Calcium (%)	0895	3.3050	0.01000	3.4502	0.17245	0.08560	141	-0.84	019.00	0
019	Calcium (%)	0878	3.3200	0.04000	3.4502	0.17245	0.08560	141	-0.75	019.00	0
019	Calcium (%)	0265	3.3250	0.35000	3.4502	0.17245	0.08560	141	-0.73	019.42	0
019	Calcium (%)	0171	3.3350	0.03000	3.4502	0.17245	0.08560	141	-0.67	019.41	0
019	Calcium (%)	0910	3.3350	0.05000	3.4502	0.17245	0.08560	141	-0.67	019.41	0
019	Calcium (%)	0298	3.3400	0.02000	3.4502	0.17245	0.08560	141	-0.64	019.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	0413	3.3500	0.10000	3.4502	0.17245	0.08560	141	-0.58	019.41	0
019	Calcium (%)	2012	3.3500	0.06000	3.4502	0.17245	0.08560	141	-0.58	019.41	0
019	Calcium (%)	0218	3.3500	0.30000	3.4502	0.17245	0.08560	141	-0.58	019.43	0
019	Calcium (%)	0013	3.3550	0.01000	3.4502	0.17245	0.08560	141	-0.55	019.43	0
019	Calcium (%)	0536	3.3650	0.19000	3.4502	0.17245	0.08560	141	-0.49	019.02	0
019	Calcium (%)	0590	3.3650	0.05000	3.4502	0.17245	0.08560	141	-0.49	019.08	0
019	Calcium (%)	0939	3.3650	0.03000	3.4502	0.17245	0.08560	141	-0.49	019.31	0
019	Calcium (%)	0169	3.3650	0.03000	3.4502	0.17245	0.08560	141	-0.49	019.32	0
019	Calcium (%)	0045	3.3650	0.17000	3.4502	0.17245	0.08560	141	-0.49	019.42	0
019	Calcium (%)	0897	3.3700	0.02000	3.4502	0.17245	0.08560	141	-0.46	019.00	0
019	Calcium (%)	0968	3.3730	0.01000	3.4502	0.17245	0.08560	141	-0.45	019.43	0
019	Calcium (%)	0089	3.3750	0.01000	3.4502	0.17245	0.08560	141	-0.44	019.43	0
019	Calcium (%)	0003	3.3800	0.08000	3.4502	0.17245	0.08560	141	-0.41	019.41	0
019	Calcium (%)	0294	3.3800	0.06000	3.4502	0.17245	0.08560	141	-0.41	019.42	0
019	Calcium (%)	0098	3.3800	0.04000	3.4502	0.17245	0.08560	141	-0.41	019.43	0
019	Calcium (%)	0693	3.3805	0.03700	3.4502	0.17245	0.08560	141	-0.40	019.42	0
019	Calcium (%)	2129	3.3815	0.19300	3.4502	0.17245	0.08560	141	-0.40	019.42	0
019	Calcium (%)	2195	3.3850	0.01000	3.4502	0.17245	0.08560	141	-0.38	019.31	0
019	Calcium (%)	0148	3.3850	0.01000	3.4502	0.17245	0.08560	141	-0.38	019.41	0
019	Calcium (%)	0186	3.4000	0.08000	3.4502	0.17245	0.08560	141	-0.29	019.42	0
019	Calcium (%)	2146	3.4050	0.09000	3.4502	0.17245	0.08560	141	-0.26	019.51	0
019	Calcium (%)	0038	3.4075	0.13700	3.4502	0.17245	0.08560	141	-0.25	019.31	0
019	Calcium (%)	0881	3.4080	0.05400	3.4502	0.17245	0.08560	141	-0.24	019.00	0
019	Calcium (%)	0622	3.4086	0.00000	3.4502	0.17245	0.08560	141	-0.24	019.00	0
019	Calcium (%)	0186	3.4100	0.04000	3.4502	0.17245	0.08560	141	-0.23	019.52	0
019	Calcium (%)	2188	3.4150	0.15000	3.4502	0.17245	0.08560	141	-0.20	019.08	0
019	Calcium (%)	0651	3.4170	0.01600	3.4502	0.17245	0.08560	141	-0.19	019.00	0
019	Calcium (%)	0849	3.4200	0.02000	3.4502	0.17245	0.08560	141	-0.17	019.08	0
019	Calcium (%)	0047	3.4200	0.04000	3.4502	0.17245	0.08560	141	-0.17	019.52	0
019	Calcium (%)	0687	3.4250	0.25000	3.4502	0.17245	0.08560	141	-0.15	019.31	0
019	Calcium (%)	0948	3.4250	0.01000	3.4502	0.17245	0.08560	141	-0.15	019.31	0
019	Calcium (%)	2188	3.4250	0.09000	3.4502	0.17245	0.08560	141	-0.15	019.31	0
019	Calcium (%)	0689	3.4300	0.02000	3.4502	0.17245	0.08560	141	-0.12	019.08	0
019	Calcium (%)	0139	3.4300	0.01400	3.4502	0.17245	0.08560	141	-0.12	019.41	0
019	Calcium (%)	0278	3.4350	0.07000	3.4502	0.17245	0.08560	141	-0.09	019.42	0
019	Calcium (%)	2109	3.4370	0.12400	3.4502	0.17245	0.08560	141	-0.08	019.00	0
019	Calcium (%)	0208	3.4395	0.04700	3.4502	0.17245	0.08560	141	-0.06	019.41	0
019	Calcium (%)	0035	3.4400	0.40000	3.4502	0.17245	0.08560	141	-0.06	019.43	0
019	Calcium (%)	0505	3.4450	0.09000	3.4502	0.17245	0.08560	141	-0.03	019.43	0
019	Calcium (%)	0668	3.4450	0.01000	3.4502	0.17245	0.08560	141	-0.03	019.43	0
019	Calcium (%)	0010	3.4500	0.02000	3.4502	0.17245	0.08560	141	0.00	019.43	0
019	Calcium (%)	0675	3.4500	0.08000	3.4502	0.17245	0.08560	141	0.00	019.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
019	Calcium (%)	0074	3.4550	0.01000	3.4502	0.17245	0.08560	141	0.03	019.41	0
019	Calcium (%)	0229	3.4555	0.08100	3.4502	0.17245	0.08560	141	0.03	019.41	0
019	Calcium (%)	0870	3.4560	0.00600	3.4502	0.17245	0.08560	141	0.03	019.42	0
019	Calcium (%)	0674	3.4600	0.00000	3.4502	0.17245	0.08560	141	0.06	019.31	0
019	Calcium (%)	0035	3.4600	0.10000	3.4502	0.17245	0.08560	141	0.06	019.42	0
019	Calcium (%)	0598	3.4615	0.15300	3.4502	0.17245	0.08560	141	0.07	019.41	0
019	Calcium (%)	0009	3.4664	0.20840	3.4502	0.17245	0.08560	141	0.09	019.42	0
019	Calcium (%)	0033	3.4700	0.12000	3.4502	0.17245	0.08560	141	0.12	019.43	0
019	Calcium (%)	0083	3.4700	0.04000	3.4502	0.17245	0.08560	141	0.12	019.43	0
019	Calcium (%)	0034	3.4795	0.11900	3.4502	0.17245	0.08560	141	0.17	019.43	0
019	Calcium (%)	0123	3.4800	0.00000	3.4502	0.17245	0.08560	141	0.17	019.41	0
019	Calcium (%)	0560	3.4827	0.12060	3.4502	0.17245	0.08560	141	0.19	019.42	0
019	Calcium (%)	2097	3.4886	0.02070	3.4502	0.17245	0.08560	141	0.22	019.31	0
019	Calcium (%)	0868	3.4900	0.02000	3.4502	0.17245	0.08560	141	0.23	019.31	0
019	Calcium (%)	0019	3.4900	0.04000	3.4502	0.17245	0.08560	141	0.23	019.41	0
019	Calcium (%)	2089	3.4900	0.14000	3.4502	0.17245	0.08560	141	0.23	019.43	0
019	Calcium (%)	0037	3.4950	0.57000	3.4502	0.17245	0.08560	141	0.26	019.42	0
019	Calcium (%)	0142	3.4967	0.14750	3.4502	0.17245	0.08560	141	0.27	019.31	0
019	Calcium (%)	0653	3.5145	0.05700	3.4502	0.17245	0.08560	141	0.37	019.41	0
019	Calcium (%)	0964	3.5231	0.11320	3.4502	0.17245	0.08560	141	0.42	019.43	0
019	Calcium (%)	0353	3.5350	0.01000	3.4502	0.17245	0.08560	141	0.49	019.43	0
019	Calcium (%)	2196	3.5400	0.00000	3.4502	0.17245	0.08560	141	0.52	019.08	0
019	Calcium (%)	2196	3.5400	0.00000	3.4502	0.17245	0.08560	141	0.52	019.31	0
019	Calcium (%)	0572	3.5400	0.20000	3.4502	0.17245	0.08560	141	0.52	019.53	0
019	Calcium (%)	0242	3.5400	0.04000	3.4502	0.17245	0.08560	141	0.52	019.99	0
019	Calcium (%)	0354	3.5450	0.01000	3.4502	0.17245	0.08560	141	0.55	019.31	0
019	Calcium (%)	0874	3.5500	0.04000	3.4502	0.17245	0.08560	141	0.58	019.31	0
019	Calcium (%)	0682	3.5500	0.00000	3.4502	0.17245	0.08560	141	0.58	019.43	0
019	Calcium (%)	2190	3.5545	0.05830	3.4502	0.17245	0.08560	141	0.60	019.08	0
019	Calcium (%)	0049	3.5550	0.19000	3.4502	0.17245	0.08560	141	0.61	019.41	0
019	Calcium (%)	0190	3.5550	0.09000	3.4502	0.17245	0.08560	141	0.61	019.42	0
019	Calcium (%)	0108	3.5600	0.02000	3.4502	0.17245	0.08560	141	0.64	019.31	0
019	Calcium (%)	0685	3.5700	0.02000	3.4502	0.17245	0.08560	141	0.69	019.42	0
019	Calcium (%)	0510	3.5750	0.03000	3.4502	0.17245	0.08560	141	0.72	019.43	0
019	Calcium (%)	0520	3.5800	0.20000	3.4502	0.17245	0.08560	141	0.75	019.41	0
019	Calcium (%)	0100	3.5800	0.02000	3.4502	0.17245	0.08560	141	0.75	019.99	0
019	Calcium (%)	2144	3.5850	0.03000	3.4502	0.17245	0.08560	141	0.78	019.99	0
019	Calcium (%)	0357	3.5958	0.03850	3.4502	0.17245	0.08560	141	0.84	019.42	0
019	Calcium (%)	0878	3.6000	0.06000	3.4502	0.17245	0.08560	141	0.87	019.41	0
019	Calcium (%)	0226	3.6000	0.04000	3.4502	0.17245	0.08560	141	0.87	019.41	0
019	Calcium (%)	0036	3.6048	0.00800	3.4502	0.17245	0.08560	141	0.90	019.03	0
019	Calcium (%)	2193	3.6200	0.00000	3.4502	0.17245	0.08560	141	0.98	019.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
019	Calcium (%)	0722	3.6253	0.03900	3.4502	0.17245	0.08560	141	1.02	019.31	0
019	Calcium (%)	0187	3.6300	0.00000	3.4502	0.17245	0.08560	141	1.04	019.42	0
019	Calcium (%)	0612	3.6350	0.27000	3.4502	0.17245	0.08560	141	1.07	019.33	0
019	Calcium (%)	2126	3.6350	0.03000	3.4502	0.17245	0.08560	141	1.07	019.99	0
019	Calcium (%)	0027	3.6435	0.17300	3.4502	0.17245	0.08560	141	1.12	019.43	0
019	Calcium (%)	0407	3.6458	0.09890	3.4502	0.17245	0.08560	141	1.13	019.41	0
019	Calcium (%)	0205	3.6600	0.06000	3.4502	0.17245	0.08560	141	1.22	019.42	0
019	Calcium (%)	0918	3.6710	0.07600	3.4502	0.17245	0.08560	141	1.28	019.43	0
019	Calcium (%)	0017	3.6765	0.10500	3.4502	0.17245	0.08560	141	1.31	019.43	0
019	Calcium (%)	0553	3.6800	0.14000	3.4502	0.17245	0.08560	141	1.33	019.53	0
019	Calcium (%)	2105	3.6900	0.24000	3.4502	0.17245	0.08560	141	1.39	019.99	0
019	Calcium (%)	0425	3.6950	0.19000	3.4502	0.17245	0.08560	141	1.42	019.43	0
019	Calcium (%)	2006	3.7200	0.00800	3.4502	0.17245	0.08560	141	1.56	019.00	0
019	Calcium (%)	0358	3.7400	0.10000	3.4502	0.17245	0.08560	141	1.68	019.41	0
019	Calcium (%)	0695	3.7400	0.02600	3.4502	0.17245	0.08560	141	1.68	019.41	0
019	Calcium (%)	2022	3.7450	0.05000	3.4502	0.17245	0.08560	141	1.71	019.31	0
019	Calcium (%)	2022	3.7650	0.03000	3.4502	0.17245	0.08560	141	1.83	019.00	0
019	Calcium (%)	0032	3.7960	0.10200	3.4502	0.17245	0.08560	141	2.01	019.43	0
019	Calcium (%)	0001	3.8965	0.06300	3.4502	0.17245	0.08560	141	2.59	019.31	0
019	Calcium (%)	0852	3.9000	0.20000	3.4502	0.17245	0.08560	141	2.61	019.08	0
019	Calcium (%)	2006	3.9020	0.00000	3.4502	0.17245	0.08560	141	2.62	019.09	0
019	Calcium (%)	0650	3.9100	0.08000	3.4502	0.17245	0.08560	141	2.67	019.31	0
019	Calcium (%)	0152	4.0000	0.18000	3.4502	0.17245	0.08560	141	3.19	019.31	0
019	Calcium (%)	0154	4.0171	0.30650	3.4502	0.17245	0.08560	141	3.29	019.52	0
019	Calcium (%)	2009	5.1314	2.9973	3.4502	0.17245	0.08560	141	9.75	019.08	1
019	Calcium (%)	0233	0.33000	0.00000	3.4502	0.17245	0.08560	141	-18.09	019.31	2
021	Cobalt (mg / kg (ppm))	0148	4.1190	0.00200	9.0953	1.6285	0.54632	26	-3.06	021.41	0
021	Cobalt (mg / kg (ppm))	0619	4.4750	0.17000	9.0953	1.6285	0.54632	26	-2.84	021.41	0
021	Cobalt (mg / kg (ppm))	0186	7.0000	0.00000	9.0953	1.6285	0.54632	26	-1.29	021.42	0
021	Cobalt (mg / kg (ppm))	0345	7.1800	0.26000	9.0953	1.6285	0.54632	26	-1.18	021.43	0
021	Cobalt (mg / kg (ppm))	0910	7.3200	0.10000	9.0953	1.6285	0.54632	26	-1.09	021.52	0
021	Cobalt (mg / kg (ppm))	0186	8.0000	0.00000	9.0953	1.6285	0.54632	26	-0.67	021.52	0
021	Cobalt (mg / kg (ppm))	0098	8.2420	0.46800	9.0953	1.6285	0.54632	26	-0.52	021.53	0
021	Cobalt (mg / kg (ppm))	0171	8.2640	0.07400	9.0953	1.6285	0.54632	26	-0.51	021.41	0
021	Cobalt (mg / kg (ppm))	2055	8.3670	2.0000	9.0953	1.6285	0.54632	26	-0.45	021.53	0
021	Cobalt (mg / kg (ppm))	0045	8.7200	0.40000	9.0953	1.6285	0.54632	26	-0.23	021.42	0
021	Cobalt (mg / kg (ppm))	0169	8.8250	0.25000	9.0953	1.6285	0.54632	26	-0.17	021.43	0
021	Cobalt (mg / kg (ppm))	0011	8.8255	0.02700	9.0953	1.6285	0.54632	26	-0.17	021.41	0
021	Cobalt (mg / kg (ppm))	0693	8.9740	1.0760	9.0953	1.6285	0.54632	26	-0.07	021.42	0
021	Cobalt (mg / kg (ppm))	0047	9.1000	0.18000	9.0953	1.6285	0.54632	26	0.00	021.52	0
021	Cobalt (mg / kg (ppm))	0964	9.2794	0.10210	9.0953	1.6285	0.54632	26	0.11	021.43	0
021	Cobalt (mg / kg (ppm))	0510	9.5700	0.04000	9.0953	1.6285	0.54632	26	0.29	021.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
021	Cobalt (mg / kg (ppm))	0560	9.7000	0.46000	9.0953	1.6285	0.54632	26	0.37	021.42	0
021	Cobalt (mg / kg (ppm))	0610	9.7600	2.0800	9.0953	1.6285	0.54632	26	0.41	021.34	0
021	Cobalt (mg / kg (ppm))	0939	10.095	0.47000	9.0953	1.6285	0.54632	26	0.61	021.31	0
021	Cobalt (mg / kg (ppm))	0918	10.226	0.57220	9.0953	1.6285	0.54632	26	0.69	021.53	0
021	Cobalt (mg / kg (ppm))	0218	10.500	1.0000	9.0953	1.6285	0.54632	26	0.86	021.43	0
021	Cobalt (mg / kg (ppm))	0038	10.510	1.2600	9.0953	1.6285	0.54632	26	0.87	021.43	0
021	Cobalt (mg / kg (ppm))	0027	10.844	1.7530	9.0953	1.6285	0.54632	26	1.07	021.43	0
021	Cobalt (mg / kg (ppm))	0553	11.000	0.40000	9.0953	1.6285	0.54632	26	1.17	021.53	0
021	Cobalt (mg / kg (ppm))	0032	11.370	0.06000	9.0953	1.6285	0.54632	26	1.40	021.53	0
021	Cobalt (mg / kg (ppm))	0218	11.500	1.0000	9.0953	1.6285	0.54632	26	1.48	021.53	0
022	Copper (mg / kg (ppm))	0868	166.18	5.6000	233.02	25.638	7.9269	101	-2.61	022.31	0
022	Copper (mg / kg (ppm))	0003	174.00	20.000	233.02	25.638	7.9269	101	-2.30	022.41	0
022	Copper (mg / kg (ppm))	0148	174.11	5.4300	233.02	25.638	7.9269	101	-2.30	022.41	0
022	Copper (mg / kg (ppm))	0910	174.50	1.0000	233.02	25.638	7.9269	101	-2.28	022.41	0
022	Copper (mg / kg (ppm))	0049	177.75	28.390	233.02	25.638	7.9269	101	-2.16	022.41	0
022	Copper (mg / kg (ppm))	0358	181.30	2.5600	233.02	25.638	7.9269	101	-2.02	022.41	0
022	Copper (mg / kg (ppm))	0619	182.50	5.0000	233.02	25.638	7.9269	101	-1.97	022.41	0
022	Copper (mg / kg (ppm))	2055	184.00	2.0000	233.02	25.638	7.9269	101	-1.91	022.53	0
022	Copper (mg / kg (ppm))	2145	184.13	3.2200	233.02	25.638	7.9269	101	-1.91	022.41	0
022	Copper (mg / kg (ppm))	0590	188.00	10.000	233.02	25.638	7.9269	101	-1.76	022.99	0
022	Copper (mg / kg (ppm))	0164	195.00	10.000	233.02	25.638	7.9269	101	-1.48	022.41	0
022	Copper (mg / kg (ppm))	0405	195.00	0.00000	233.02	25.638	7.9269	101	-1.48	022.41	0
022	Copper (mg / kg (ppm))	0098	195.85	22.900	233.02	25.638	7.9269	101	-1.45	022.41	0
022	Copper (mg / kg (ppm))	2012	198.34	6.1100	233.02	25.638	7.9269	101	-1.35	022.41	0
022	Copper (mg / kg (ppm))	0037	199.00	8.0000	233.02	25.638	7.9269	101	-1.33	022.42	0
022	Copper (mg / kg (ppm))	0407	201.37	0.07460	233.02	25.638	7.9269	101	-1.23	022.41	0
022	Copper (mg / kg (ppm))	0038	202.00	0.00000	233.02	25.638	7.9269	101	-1.21	022.32	0
022	Copper (mg / kg (ppm))	2146	204.40	4.0000	233.02	25.638	7.9269	101	-1.12	022.51	0
022	Copper (mg / kg (ppm))	0171	206.05	2.1000	233.02	25.638	7.9269	101	-1.05	022.41	0
022	Copper (mg / kg (ppm))	0520	207.50	23.000	233.02	25.638	7.9269	101	-1.00	022.41	0
022	Copper (mg / kg (ppm))	0354	209.65	2.2300	233.02	25.638	7.9269	101	-0.91	022.31	0
022	Copper (mg / kg (ppm))	0175	210.00	0.00000	233.02	25.638	7.9269	101	-0.90	022.31	0
022	Copper (mg / kg (ppm))	0878	210.00	2.0000	233.02	25.638	7.9269	101	-0.90	022.41	0
022	Copper (mg / kg (ppm))	0011	212.29	0.25700	233.02	25.638	7.9269	101	-0.81	022.41	0
022	Copper (mg / kg (ppm))	2097	213.49	2.2787	233.02	25.638	7.9269	101	-0.76	022.31	0
022	Copper (mg / kg (ppm))	0921	214.44	31.650	233.02	25.638	7.9269	101	-0.73	022.31	0
022	Copper (mg / kg (ppm))	2192	214.69	4.8100	233.02	25.638	7.9269	101	-0.72	022.43	0
022	Copper (mg / kg (ppm))	0001	215.20	7.2000	233.02	25.638	7.9269	101	-0.70	022.31	0
022	Copper (mg / kg (ppm))	0345	219.00	8.0000	233.02	25.638	7.9269	101	-0.55	022.43	0
022	Copper (mg / kg (ppm))	0511	219.50	11.000	233.02	25.638	7.9269	101	-0.53	022.41	0
022	Copper (mg / kg (ppm))	0536	221.56	2.9500	233.02	25.638	7.9269	101	-0.45	022.31	0
022	Copper (mg / kg (ppm))	0229	221.80	8.3800	233.02	25.638	7.9269	101	-0.44	022.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
022	Copper (mg / kg (ppm))	2195	221.97	1.6800	233.02	25.638	7.9269	101	-0.43	022.31	0
022	Copper (mg / kg (ppm))	0019	223.20	9.8000	233.02	25.638	7.9269	101	-0.38	022.41	0
022	Copper (mg / kg (ppm))	0598	226.55	2.1000	233.02	25.638	7.9269	101	-0.25	022.41	0
022	Copper (mg / kg (ppm))	0948	227.10	1.0550	233.02	25.638	7.9269	101	-0.23	022.33	0
022	Copper (mg / kg (ppm))	0208	227.50	23.000	233.02	25.638	7.9269	101	-0.22	022.31	0
022	Copper (mg / kg (ppm))	0860	228.63	0.00000	233.02	25.638	7.9269	101	-0.17	022.43	0
022	Copper (mg / kg (ppm))	0208	229.70	6.4000	233.02	25.638	7.9269	101	-0.13	022.41	0
022	Copper (mg / kg (ppm))	0074	232.50	5.0000	233.02	25.638	7.9269	101	-0.02	022.41	0
022	Copper (mg / kg (ppm))	2188	232.56	10.340	233.02	25.638	7.9269	101	-0.02	022.31	0
022	Copper (mg / kg (ppm))	0265	233.50	1.0000	233.02	25.638	7.9269	101	0.02	022.42	0
022	Copper (mg / kg (ppm))	2022	233.66	0.33000	233.02	25.638	7.9269	101	0.02	022.31	0
022	Copper (mg / kg (ppm))	2196	233.70	0.00000	233.02	25.638	7.9269	101	0.03	022.31	0
022	Copper (mg / kg (ppm))	0202	234.25	9.7000	233.02	25.638	7.9269	101	0.05	022.43	0
022	Copper (mg / kg (ppm))	0218	235.00	10.000	233.02	25.638	7.9269	101	0.08	022.43	0
022	Copper (mg / kg (ppm))	2128	235.94	22.549	233.02	25.638	7.9269	101	0.11	022.32	0
022	Copper (mg / kg (ppm))	0226	236.00	30.000	233.02	25.638	7.9269	101	0.12	022.41	0
022	Copper (mg / kg (ppm))	0968	236.00	0.00000	233.02	25.638	7.9269	101	0.12	022.43	0
022	Copper (mg / kg (ppm))	0278	236.15	4.3000	233.02	25.638	7.9269	101	0.12	022.42	0
022	Copper (mg / kg (ppm))	0939	236.88	8.4700	233.02	25.638	7.9269	101	0.15	022.31	0
022	Copper (mg / kg (ppm))	0294	238.63	3.8600	233.02	25.638	7.9269	101	0.22	022.42	0
022	Copper (mg / kg (ppm))	0035	240.50	9.0000	233.02	25.638	7.9269	101	0.29	022.43	0
022	Copper (mg / kg (ppm))	0098	240.70	13.800	233.02	25.638	7.9269	101	0.30	022.43	0
022	Copper (mg / kg (ppm))	0013	241.00	20.000	233.02	25.638	7.9269	101	0.31	022.43	0
022	Copper (mg / kg (ppm))	0169	241.00	8.0000	233.02	25.638	7.9269	101	0.31	022.43	0
022	Copper (mg / kg (ppm))	0572	241.00	4.0000	233.02	25.638	7.9269	101	0.31	022.53	0
022	Copper (mg / kg (ppm))	0918	241.51	6.9700	233.02	25.638	7.9269	101	0.33	022.43	0
022	Copper (mg / kg (ppm))	0553	242.00	2.0000	233.02	25.638	7.9269	101	0.35	022.53	0
022	Copper (mg / kg (ppm))	2144	242.46	1.5500	233.02	25.638	7.9269	101	0.37	022.99	0
022	Copper (mg / kg (ppm))	0964	246.16	23.210	233.02	25.638	7.9269	101	0.51	022.43	0
022	Copper (mg / kg (ppm))	0100	246.50	3.0000	233.02	25.638	7.9269	101	0.53	022.99	0
022	Copper (mg / kg (ppm))	0045	247.00	16.000	233.02	25.638	7.9269	101	0.55	022.42	0
022	Copper (mg / kg (ppm))	0242	247.50	5.0000	233.02	25.638	7.9269	101	0.56	022.99	0
022	Copper (mg / kg (ppm))	0722	248.21	2.5864	233.02	25.638	7.9269	101	0.59	022.31	0
022	Copper (mg / kg (ppm))	0425	248.75	0.70000	233.02	25.638	7.9269	101	0.61	022.43	0
022	Copper (mg / kg (ppm))	0870	249.05	1.7000	233.02	25.638	7.9269	101	0.63	022.42	0
022	Copper (mg / kg (ppm))	0083	249.50	7.0000	233.02	25.638	7.9269	101	0.64	022.43	0
022	Copper (mg / kg (ppm))	0047	249.66	0.28000	233.02	25.638	7.9269	101	0.65	022.52	0
022	Copper (mg / kg (ppm))	0560	249.80	8.0000	233.02	25.638	7.9269	101	0.65	022.42	0
022	Copper (mg / kg (ppm))	0190	249.90	12.610	233.02	25.638	7.9269	101	0.66	022.42	0
022	Copper (mg / kg (ppm))	0033	250.00	4.0000	233.02	25.638	7.9269	101	0.66	022.43	0
022	Copper (mg / kg (ppm))	0861	250.00	18.000	233.02	25.638	7.9269	101	0.66	022.43	0
022	Copper (mg / kg (ppm))	0413	250.50	5.0000	233.02	25.638	7.9269	101	0.68	022.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
022	Copper (mg / kg (ppm))	0693	250.94	13.780	233.02	25.638	7.9269	101	0.70	022.42	0
022	Copper (mg / kg (ppm))	0186	251.00	4.0000	233.02	25.638	7.9269	101	0.70	022.52	0
022	Copper (mg / kg (ppm))	0357	251.26	0.19500	233.02	25.638	7.9269	101	0.71	022.42	0
022	Copper (mg / kg (ppm))	0682	251.30	0.00000	233.02	25.638	7.9269	101	0.71	022.43	0
022	Copper (mg / kg (ppm))	0038	251.73	7.8400	233.02	25.638	7.9269	101	0.73	022.43	0
022	Copper (mg / kg (ppm))	0510	252.50	1.0000	233.02	25.638	7.9269	101	0.76	022.43	0
022	Copper (mg / kg (ppm))	0187	253.80	3.0000	233.02	25.638	7.9269	101	0.81	022.42	0
022	Copper (mg / kg (ppm))	2129	253.95	20.900	233.02	25.638	7.9269	101	0.82	022.99	0
022	Copper (mg / kg (ppm))	0668	254.00	20.000	233.02	25.638	7.9269	101	0.82	022.43	0
022	Copper (mg / kg (ppm))	0009	254.50	21.000	233.02	25.638	7.9269	101	0.84	022.42	0
022	Copper (mg / kg (ppm))	0032	254.75	5.5000	233.02	25.638	7.9269	101	0.85	022.43	0
022	Copper (mg / kg (ppm))	0689	255.05	18.100	233.02	25.638	7.9269	101	0.86	022.31	0
022	Copper (mg / kg (ppm))	0675	255.45	6.8600	233.02	25.638	7.9269	101	0.87	022.43	0
022	Copper (mg / kg (ppm))	0035	255.50	1.0000	233.02	25.638	7.9269	101	0.88	022.42	0
022	Copper (mg / kg (ppm))	0034	255.85	19.100	233.02	25.638	7.9269	101	0.89	022.43	0
022	Copper (mg / kg (ppm))	0353	256.00	2.0000	233.02	25.638	7.9269	101	0.90	022.43	0
022	Copper (mg / kg (ppm))	0010	257.90	5.0000	233.02	25.638	7.9269	101	0.97	022.33	0
022	Copper (mg / kg (ppm))	2053	258.00	12.000	233.02	25.638	7.9269	101	0.97	022.42	0
022	Copper (mg / kg (ppm))	0610	258.00	6.0000	233.02	25.638	7.9269	101	0.97	022.43	0
022	Copper (mg / kg (ppm))	0218	260.00	0.00000	233.02	25.638	7.9269	101	1.05	022.53	0
022	Copper (mg / kg (ppm))	0027	261.73	24.970	233.02	25.638	7.9269	101	1.12	022.43	0
022	Copper (mg / kg (ppm))	0042	262.00	18.000	233.02	25.638	7.9269	101	1.13	022.42	0
022	Copper (mg / kg (ppm))	0186	264.00	6.0000	233.02	25.638	7.9269	101	1.21	022.42	0
022	Copper (mg / kg (ppm))	0205	264.50	7.0000	233.02	25.638	7.9269	101	1.23	022.42	0
022	Copper (mg / kg (ppm))	0505	268.15	5.1000	233.02	25.638	7.9269	101	1.37	022.43	0
022	Copper (mg / kg (ppm))	0685	268.29	1.0700	233.02	25.638	7.9269	101	1.38	022.42	0
022	Copper (mg / kg (ppm))	0017	274.39	9.0690	233.02	25.638	7.9269	101	1.61	022.43	0
022	Copper (mg / kg (ppm))	2089	245.05	75.830	233.02	25.638	7.9269	101	0.47	022.43	1
023	Fluorine (mg / kg (ppm))	0910	53.600	4.2000				1		023.01	0
024	Iodine (mg / kg (ppm))	0160	16.900	1.4000				1		024.99	0
025	Iron (mg / kg (ppm))	0278	338.50	5.0000	883.57	65.105	36.603	78	-8.37	025.42	0
025	Iron (mg / kg (ppm))	0265	411.50	81.000	883.57	65.105	36.603	78	-7.25	025.42	0
025	Iron (mg / kg (ppm))	0619	625.50	13.000	883.57	65.105	36.603	78	-3.96	025.41	0
025	Iron (mg / kg (ppm))	0187	642.82	24.570	883.57	65.105	36.603	78	-3.70	025.42	0
025	Iron (mg / kg (ppm))	0226	722.50	55.000	883.57	65.105	36.603	78	-2.47	025.41	0
025	Iron (mg / kg (ppm))	2012	731.99	17.320	883.57	65.105	36.603	78	-2.33	025.41	0
025	Iron (mg / kg (ppm))	0693	766.67	167.92	883.57	65.105	36.603	78	-1.80	025.42	0
025	Iron (mg / kg (ppm))	2128	769.89	201.08	883.57	65.105	36.603	78	-1.75	025.32	0
025	Iron (mg / kg (ppm))	0294	779.72	1.4400	883.57	65.105	36.603	78	-1.60	025.42	0
025	Iron (mg / kg (ppm))	0968	796.50	3.0000	883.57	65.105	36.603	78	-1.34	025.43	0
025	Iron (mg / kg (ppm))	0689	797.95	14.300	883.57	65.105	36.603	78	-1.32	025.31	0
025	Iron (mg / kg (ppm))	0353	802.00	26.000	883.57	65.105	36.603	78	-1.25	025.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
025	Iron (mg / kg (ppm))	0870	809.25	0.70000	883.57	65.105	36.603	78	-1.14	025.42	0
025	Iron (mg / kg (ppm))	0598	812.15	25.700	883.57	65.105	36.603	78	-1.10	025.41	0
025	Iron (mg / kg (ppm))	0560	817.85	20.700	883.57	65.105	36.603	78	-1.01	025.42	0
025	Iron (mg / kg (ppm))	0511	822.50	95.000	883.57	65.105	36.603	78	-0.94	025.41	0
025	Iron (mg / kg (ppm))	0037	827.50	19.000	883.57	65.105	36.603	78	-0.86	025.42	0
025	Iron (mg / kg (ppm))	0169	833.50	23.000	883.57	65.105	36.603	78	-0.77	025.43	0
025	Iron (mg / kg (ppm))	2097	835.43	7.0250	883.57	65.105	36.603	78	-0.74	025.31	0
025	Iron (mg / kg (ppm))	2192	837.86	14.300	883.57	65.105	36.603	78	-0.70	025.43	0
025	Iron (mg / kg (ppm))	2195	839.04	27.020	883.57	65.105	36.603	78	-0.68	025.31	0
025	Iron (mg / kg (ppm))	0407	842.06	53.649	883.57	65.105	36.603	78	-0.64	025.41	0
025	Iron (mg / kg (ppm))	0682	847.23	0.00000	883.57	65.105	36.603	78	-0.56	025.43	0
025	Iron (mg / kg (ppm))	0405	848.00	0.00000	883.57	65.105	36.603	78	-0.55	025.41	0
025	Iron (mg / kg (ppm))	0003	857.00	48.000	883.57	65.105	36.603	78	-0.41	025.41	0
025	Iron (mg / kg (ppm))	0878	861.50	9.0000	883.57	65.105	36.603	78	-0.34	025.41	0
025	Iron (mg / kg (ppm))	0190	861.52	29.990	883.57	65.105	36.603	78	-0.34	025.42	0
025	Iron (mg / kg (ppm))	2145	862.46	8.3600	883.57	65.105	36.603	78	-0.32	025.41	0
025	Iron (mg / kg (ppm))	0164	868.00	34.000	883.57	65.105	36.603	78	-0.24	025.41	0
025	Iron (mg / kg (ppm))	0413	872.50	17.000	883.57	65.105	36.603	78	-0.17	025.42	0
025	Iron (mg / kg (ppm))	0035	878.00	130.00	883.57	65.105	36.603	78	-0.09	025.43	0
025	Iron (mg / kg (ppm))	0425	879.50	13.000	883.57	65.105	36.603	78	-0.06	025.43	0
025	Iron (mg / kg (ppm))	0861	882.00	62.000	883.57	65.105	36.603	78	-0.02	025.43	0
025	Iron (mg / kg (ppm))	0520	883.50	45.000	883.57	65.105	36.603	78	0.00	025.41	0
025	Iron (mg / kg (ppm))	0874	885.00	28.000	883.57	65.105	36.603	78	0.02	025.31	0
025	Iron (mg / kg (ppm))	0045	885.00	10.000	883.57	65.105	36.603	78	0.02	025.42	0
025	Iron (mg / kg (ppm))	0017	886.89	34.931	883.57	65.105	36.603	78	0.05	025.43	0
025	Iron (mg / kg (ppm))	0098	890.50	18.800	883.57	65.105	36.603	78	0.11	025.41	0
025	Iron (mg / kg (ppm))	0208	892.95	19.500	883.57	65.105	36.603	78	0.14	025.41	0
025	Iron (mg / kg (ppm))	0910	897.50	3.0000	883.57	65.105	36.603	78	0.21	025.41	0
025	Iron (mg / kg (ppm))	0083	899.00	24.000	883.57	65.105	36.603	78	0.24	025.43	0
025	Iron (mg / kg (ppm))	0175	900.00	0.00000	883.57	65.105	36.603	78	0.25	025.31	0
025	Iron (mg / kg (ppm))	0171	904.38	17.250	883.57	65.105	36.603	78	0.32	025.41	0
025	Iron (mg / kg (ppm))	0242	905.50	27.000	883.57	65.105	36.603	78	0.34	025.99	0
025	Iron (mg / kg (ppm))	0019	906.25	78.900	883.57	65.105	36.603	78	0.35	025.41	0
025	Iron (mg / kg (ppm))	2196	907.70	0.00000	883.57	65.105	36.603	78	0.37	025.31	0
025	Iron (mg / kg (ppm))	0918	908.30	19.110	883.57	65.105	36.603	78	0.38	025.43	0
025	Iron (mg / kg (ppm))	0100	912.50	5.0000	883.57	65.105	36.603	78	0.44	025.99	0
025	Iron (mg / kg (ppm))	0074	916.50	17.000	883.57	65.105	36.603	78	0.51	025.41	0
025	Iron (mg / kg (ppm))	0049	916.81	48.990	883.57	65.105	36.603	78	0.51	025.41	0
025	Iron (mg / kg (ppm))	0675	918.34	18.720	883.57	65.105	36.603	78	0.53	025.43	0
025	Iron (mg / kg (ppm))	0042	918.50	19.000	883.57	65.105	36.603	78	0.54	025.42	0
025	Iron (mg / kg (ppm))	2188	919.00	212.00	883.57	65.105	36.603	78	0.54	025.31	0
025	Iron (mg / kg (ppm))	0536	920.57	34.130	883.57	65.105	36.603	78	0.57	025.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
025	Iron (mg / kg (ppm))	0038	921.00	12.000	883.57	65.105	36.603	78	0.57	025.31	0
025	Iron (mg / kg (ppm))	2053	921.00	44.000	883.57	65.105	36.603	78	0.57	025.42	0
025	Iron (mg / kg (ppm))	0098	921.15	16.900	883.57	65.105	36.603	78	0.58	025.43	0
025	Iron (mg / kg (ppm))	0964	927.89	61.982	883.57	65.105	36.603	78	0.68	025.43	0
025	Iron (mg / kg (ppm))	0358	929.97	28.760	883.57	65.105	36.603	78	0.71	025.41	0
025	Iron (mg / kg (ppm))	0218	930.00	40.000	883.57	65.105	36.603	78	0.71	025.43	0
025	Iron (mg / kg (ppm))	0939	931.02	26.740	883.57	65.105	36.603	78	0.73	025.31	0
025	Iron (mg / kg (ppm))	0722	931.55	12.769	883.57	65.105	36.603	78	0.74	025.31	0
025	Iron (mg / kg (ppm))	0354	938.11	4.4400	883.57	65.105	36.603	78	0.84	025.31	0
025	Iron (mg / kg (ppm))	0553	938.50	33.000	883.57	65.105	36.603	78	0.84	025.53	0
025	Iron (mg / kg (ppm))	0868	941.28	92.140	883.57	65.105	36.603	78	0.89	025.31	0
025	Iron (mg / kg (ppm))	2146	944.40	77.200	883.57	65.105	36.603	78	0.93	025.51	0
025	Iron (mg / kg (ppm))	0148	945.97	49.160	883.57	65.105	36.603	78	0.96	025.41	0
025	Iron (mg / kg (ppm))	0610	946.00	36.000	883.57	65.105	36.603	78	0.96	025.43	0
025	Iron (mg / kg (ppm))	2022	948.89	2.7600	883.57	65.105	36.603	78	1.00	025.31	0
025	Iron (mg / kg (ppm))	0034	950.00	37.800	883.57	65.105	36.603	78	1.02	025.43	0
025	Iron (mg / kg (ppm))	0510	953.50	1.0000	883.57	65.105	36.603	78	1.07	025.43	0
025	Iron (mg / kg (ppm))	0035	954.50	15.000	883.57	65.105	36.603	78	1.09	025.42	0
025	Iron (mg / kg (ppm))	0208	955.50	109.00	883.57	65.105	36.603	78	1.10	025.31	0
025	Iron (mg / kg (ppm))	0229	957.32	56.750	883.57	65.105	36.603	78	1.13	025.41	0
025	Iron (mg / kg (ppm))	2144	961.84	4.2900	883.57	65.105	36.603	78	1.20	025.99	0
025	Iron (mg / kg (ppm))	0038	967.35	19.290	883.57	65.105	36.603	78	1.29	025.43	0
025	Iron (mg / kg (ppm))	0032	1,012.0	2.0000	883.57	65.105	36.603	78	1.97	025.43	0
025	Iron (mg / kg (ppm))	2089	1,204.3	144.67	883.57	65.105	36.603	78	4.93	025.43	0
025	Iron (mg / kg (ppm))	2055	540.00	530.00	883.57	65.105	36.603	78	-5.28	025.53	1
027	Magnesium (%)	0089	0.45500	0.01000	0.54997	0.02746	0.01271	94	-3.46	027.43	0
027	Magnesium (%)	0169	0.45950	0.00100	0.54997	0.02746	0.01271	94	-3.29	027.32	0
027	Magnesium (%)	0407	0.48705	0.02850	0.54997	0.02746	0.01271	94	-2.29	027.41	0
027	Magnesium (%)	2145	0.49190	0.00120	0.54997	0.02746	0.01271	94	-2.11	027.41	0
027	Magnesium (%)	0042	0.49450	0.01900	0.54997	0.02746	0.01271	94	-2.02	027.42	0
027	Magnesium (%)	2055	0.49670	0.03140	0.54997	0.02746	0.01271	94	-1.94	027.53	0
027	Magnesium (%)	0345	0.49700	0.01200	0.54997	0.02746	0.01271	94	-1.93	027.43	0
027	Magnesium (%)	0668	0.50550	0.01900	0.54997	0.02746	0.01271	94	-1.62	027.43	0
027	Magnesium (%)	2192	0.51000	0.00000	0.54997	0.02746	0.01271	94	-1.46	027.43	0
027	Magnesium (%)	0870	0.51060	0.00900	0.54997	0.02746	0.01271	94	-1.43	027.42	0
027	Magnesium (%)	0511	0.52000	0.02000	0.54997	0.02746	0.01271	94	-1.09	027.41	0
027	Magnesium (%)	0278	0.52000	0.02000	0.54997	0.02746	0.01271	94	-1.09	027.42	0
027	Magnesium (%)	0035	0.52000	0.02000	0.54997	0.02746	0.01271	94	-1.09	027.43	0
027	Magnesium (%)	0968	0.52050	0.00100	0.54997	0.02746	0.01271	94	-1.07	027.43	0
027	Magnesium (%)	0202	0.52200	0.01000	0.54997	0.02746	0.01271	94	-1.02	027.43	0
027	Magnesium (%)	0861	0.52600	0.02600	0.54997	0.02746	0.01271	94	-0.87	027.43	0
027	Magnesium (%)	0047	0.52700	0.00200	0.54997	0.02746	0.01271	94	-0.84	027.52	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
027	Magnesium (%)	0011	0.52725	0.00050	0.54997	0.02746	0.01271	94	-0.83	027.41	0
027	Magnesium (%)	0590	0.52900	0.01200	0.54997	0.02746	0.01271	94	-0.76	027.99	0
027	Magnesium (%)	0098	0.53000	0.00000	0.54997	0.02746	0.01271	94	-0.73	027.41	0
027	Magnesium (%)	2012	0.53000	0.00000	0.54997	0.02746	0.01271	94	-0.73	027.41	0
027	Magnesium (%)	0510	0.53000	0.00000	0.54997	0.02746	0.01271	94	-0.73	027.43	0
027	Magnesium (%)	0142	0.53215	0.00750	0.54997	0.02746	0.01271	94	-0.65	027.31	0
027	Magnesium (%)	0619	0.53250	0.00300	0.54997	0.02746	0.01271	94	-0.64	027.41	0
027	Magnesium (%)	0689	0.53500	0.01000	0.54997	0.02746	0.01271	94	-0.55	027.31	0
027	Magnesium (%)	0148	0.53500	0.01000	0.54997	0.02746	0.01271	94	-0.55	027.41	0
027	Magnesium (%)	0413	0.53500	0.01000	0.54997	0.02746	0.01271	94	-0.55	027.41	0
027	Magnesium (%)	0910	0.53500	0.01000	0.54997	0.02746	0.01271	94	-0.55	027.41	0
027	Magnesium (%)	0294	0.53500	0.01000	0.54997	0.02746	0.01271	94	-0.55	027.42	0
027	Magnesium (%)	0218	0.53500	0.01000	0.54997	0.02746	0.01271	94	-0.55	027.43	0
027	Magnesium (%)	0186	0.53500	0.00600	0.54997	0.02746	0.01271	94	-0.55	027.52	0
027	Magnesium (%)	0164	0.53700	0.00400	0.54997	0.02746	0.01271	94	-0.47	027.41	0
027	Magnesium (%)	0598	0.53790	0.00980	0.54997	0.02746	0.01271	94	-0.44	027.41	0
027	Magnesium (%)	0918	0.53850	0.01500	0.54997	0.02746	0.01271	94	-0.42	027.43	0
027	Magnesium (%)	0354	0.54000	0.00000	0.54997	0.02746	0.01271	94	-0.36	027.31	0
027	Magnesium (%)	2196	0.54000	0.00000	0.54997	0.02746	0.01271	94	-0.36	027.31	0
027	Magnesium (%)	0003	0.54000	0.02000	0.54997	0.02746	0.01271	94	-0.36	027.41	0
027	Magnesium (%)	0045	0.54050	0.06100	0.54997	0.02746	0.01271	94	-0.35	027.42	0
027	Magnesium (%)	0229	0.54350	0.01100	0.54997	0.02746	0.01271	94	-0.24	027.41	0
027	Magnesium (%)	0098	0.54500	0.05000	0.54997	0.02746	0.01271	94	-0.18	027.43	0
027	Magnesium (%)	0175	0.54500	0.01000	0.54997	0.02746	0.01271	94	-0.18	027.31	0
027	Magnesium (%)	0265	0.54500	0.03000	0.54997	0.02746	0.01271	94	-0.18	027.42	0
027	Magnesium (%)	0693	0.54550	0.01100	0.54997	0.02746	0.01271	94	-0.16	027.42	0
027	Magnesium (%)	0939	0.55000	0.02000	0.54997	0.02746	0.01271	94	0.00	027.31	0
027	Magnesium (%)	0049	0.55000	0.04000	0.54997	0.02746	0.01271	94	0.00	027.41	0
027	Magnesium (%)	0035	0.55000	0.00000	0.54997	0.02746	0.01271	94	0.00	027.42	0
027	Magnesium (%)	0083	0.55000	0.00000	0.54997	0.02746	0.01271	94	0.00	027.43	0
027	Magnesium (%)	0171	0.55050	0.00100	0.54997	0.02746	0.01271	94	0.02	027.41	0
027	Magnesium (%)	0186	0.55050	0.00300	0.54997	0.02746	0.01271	94	0.02	027.42	0
027	Magnesium (%)	0208	0.55150	0.04300	0.54997	0.02746	0.01271	94	0.06	027.31	0
027	Magnesium (%)	0032	0.55325	0.02230	0.54997	0.02746	0.01271	94	0.12	027.43	0
027	Magnesium (%)	0009	0.55435	0.03950	0.54997	0.02746	0.01271	94	0.16	027.42	0
027	Magnesium (%)	2146	0.55500	0.01000	0.54997	0.02746	0.01271	94	0.18	027.51	0
027	Magnesium (%)	0560	0.55585	0.02050	0.54997	0.02746	0.01271	94	0.21	027.42	0
027	Magnesium (%)	2129	0.55720	0.01680	0.54997	0.02746	0.01271	94	0.26	027.42	0
027	Magnesium (%)	0038	0.55725	0.04150	0.54997	0.02746	0.01271	94	0.26	027.43	0
027	Magnesium (%)	0038	0.55855	0.01230	0.54997	0.02746	0.01271	94	0.31	027.31	0
027	Magnesium (%)	0612	0.55900	0.00600	0.54997	0.02746	0.01271	94	0.33	027.33	0
027	Magnesium (%)	0405	0.56000	0.00000	0.54997	0.02746	0.01271	94	0.37	027.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
027	Magnesium (%)	0878	0.56000	0.04000	0.54997	0.02746	0.01271	94	0.37	027.41	0
027	Magnesium (%)	0013	0.56000	0.00000	0.54997	0.02746	0.01271	94	0.37	027.43	0
027	Magnesium (%)	0610	0.56000	0.02000	0.54997	0.02746	0.01271	94	0.37	027.43	0
027	Magnesium (%)	0074	0.56050	0.00500	0.54997	0.02746	0.01271	94	0.38	027.41	0
027	Magnesium (%)	0139	0.56065	0.00130	0.54997	0.02746	0.01271	94	0.39	027.41	0
027	Magnesium (%)	0034	0.56290	0.01580	0.54997	0.02746	0.01271	94	0.47	027.43	0
027	Magnesium (%)	0948	0.56300	0.00200	0.54997	0.02746	0.01271	94	0.47	027.33	0
027	Magnesium (%)	0505	0.56300	0.00400	0.54997	0.02746	0.01271	94	0.47	027.43	0
027	Magnesium (%)	0033	0.56500	0.01000	0.54997	0.02746	0.01271	94	0.55	027.43	0
027	Magnesium (%)	0242	0.56500	0.01000	0.54997	0.02746	0.01271	94	0.55	027.99	0
027	Magnesium (%)	0964	0.56610	0.03560	0.54997	0.02746	0.01271	94	0.59	027.43	0
027	Magnesium (%)	0187	0.56740	0.00200	0.54997	0.02746	0.01271	94	0.63	027.42	0
027	Magnesium (%)	0357	0.56750	0.00540	0.54997	0.02746	0.01271	94	0.64	027.42	0
027	Magnesium (%)	0868	0.57000	0.00000	0.54997	0.02746	0.01271	94	0.73	027.31	0
027	Magnesium (%)	0572	0.57300	0.00600	0.54997	0.02746	0.01271	94	0.84	027.53	0
027	Magnesium (%)	0019	0.57500	0.01000	0.54997	0.02746	0.01271	94	0.91	027.41	0
027	Magnesium (%)	0190	0.57500	0.01000	0.54997	0.02746	0.01271	94	0.91	027.42	0
027	Magnesium (%)	0425	0.57500	0.01000	0.54997	0.02746	0.01271	94	0.91	027.43	0
027	Magnesium (%)	0100	0.57500	0.01000	0.54997	0.02746	0.01271	94	0.91	027.99	0
027	Magnesium (%)	0001	0.57755	0.01430	0.54997	0.02746	0.01271	94	1.00	027.31	0
027	Magnesium (%)	0874	0.57900	0.00000	0.54997	0.02746	0.01271	94	1.06	027.31	0
027	Magnesium (%)	0226	0.58000	0.00000	0.54997	0.02746	0.01271	94	1.09	027.41	0
027	Magnesium (%)	0685	0.58000	0.00000	0.54997	0.02746	0.01271	94	1.09	027.42	0
027	Magnesium (%)	0682	0.58000	0.00000	0.54997	0.02746	0.01271	94	1.09	027.43	0
027	Magnesium (%)	0208	0.58050	0.00100	0.54997	0.02746	0.01271	94	1.11	027.41	0
027	Magnesium (%)	0553	0.58200	0.01200	0.54997	0.02746	0.01271	94	1.17	027.53	0
027	Magnesium (%)	0027	0.58200	0.05200	0.54997	0.02746	0.01271	94	1.17	027.43	0
027	Magnesium (%)	0358	0.59500	0.01000	0.54997	0.02746	0.01271	94	1.64	027.41	0
027	Magnesium (%)	0650	0.59900	0.00600	0.54997	0.02746	0.01271	94	1.79	027.31	0
027	Magnesium (%)	0675	0.60500	0.01000	0.54997	0.02746	0.01271	94	2.00	027.43	0
027	Magnesium (%)	2144	0.60950	0.01300	0.54997	0.02746	0.01271	94	2.17	027.99	0
027	Magnesium (%)	0154	0.61755	0.01370	0.54997	0.02746	0.01271	94	2.46	027.52	0
027	Magnesium (%)	0017	0.62650	0.02900	0.54997	0.02746	0.01271	94	2.79	027.43	0
027	Magnesium (%)	2089	0.63000	0.02000	0.54997	0.02746	0.01271	94	2.91	027.43	0
027	Magnesium (%)	0889	0.71590	0.00000	0.54997	0.02746	0.01271	94	6.04	027.99	0
027	Magnesium (%)	0353	0.41000	0.08000	0.54997	0.02746	0.01271	94	-5.10	027.43	1
027	Magnesium (%)	0037	0.56500	0.09000	0.54997	0.02746	0.01271	94	0.55	027.42	1
027	Magnesium (%)	0520	0.59500	0.11000	0.54997	0.02746	0.01271	94	1.64	027.41	1
027	Magnesium (%)	0536	0.27000	0.02000	0.54997	0.02746	0.01271	94	-10.20	027.31	2
028	Manganese (mg / kg (ppm))	0619	386.00	10.000	533.91	28.740	13.776	94	-5.15	028.41	0
028	Manganese (mg / kg (ppm))	0226	418.00	36.000	533.91	28.740	13.776	94	-4.03	028.41	0
028	Manganese (mg / kg (ppm))	2195	447.01	6.5500	533.91	28.740	13.776	94	-3.02	028.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
028	Manganese (mg / kg (ppm))	2055	466.45	7.1000	533.91	28.740	13.776	94	-2.35	028.53	0
028	Manganese (mg / kg (ppm))	2146	466.85	24.700	533.91	28.740	13.776	94	-2.33	028.51	0
028	Manganese (mg / kg (ppm))	0511	471.50	43.000	533.91	28.740	13.776	94	-2.17	028.41	0
028	Manganese (mg / kg (ppm))	0164	478.50	7.0000	533.91	28.740	13.776	94	-1.93	028.41	0
028	Manganese (mg / kg (ppm))	0685	480.73	1.4100	533.91	28.740	13.776	94	-1.85	028.42	0
028	Manganese (mg / kg (ppm))	0590	483.50	1.0000	533.91	28.740	13.776	94	-1.75	028.99	0
028	Manganese (mg / kg (ppm))	0037	496.00	0.00000	533.91	28.740	13.776	94	-1.32	028.42	0
028	Manganese (mg / kg (ppm))	0536	497.91	1.4500	533.91	28.740	13.776	94	-1.25	028.31	0
028	Manganese (mg / kg (ppm))	2012	499.50	4.1000	533.91	28.740	13.776	94	-1.20	028.41	0
028	Manganese (mg / kg (ppm))	0407	500.12	8.9233	533.91	28.740	13.776	94	-1.18	028.41	0
028	Manganese (mg / kg (ppm))	0968	500.50	1.0000	533.91	28.740	13.776	94	-1.16	028.43	0
028	Manganese (mg / kg (ppm))	0345	502.50	5.0000	533.91	28.740	13.776	94	-1.09	028.43	0
028	Manganese (mg / kg (ppm))	0861	503.00	42.000	533.91	28.740	13.776	94	-1.08	028.43	0
028	Manganese (mg / kg (ppm))	0098	508.45	2.7000	533.91	28.740	13.776	94	-0.89	028.41	0
028	Manganese (mg / kg (ppm))	0011	509.34	0.62200	533.91	28.740	13.776	94	-0.85	028.41	0
028	Manganese (mg / kg (ppm))	0278	509.50	7.0000	533.91	28.740	13.776	94	-0.85	028.42	0
028	Manganese (mg / kg (ppm))	0049	509.52	26.400	533.91	28.740	13.776	94	-0.85	028.41	0
028	Manganese (mg / kg (ppm))	0171	509.95	1.9000	533.91	28.740	13.776	94	-0.83	028.41	0
028	Manganese (mg / kg (ppm))	0520	510.50	45.000	533.91	28.740	13.776	94	-0.81	028.41	0
028	Manganese (mg / kg (ppm))	0265	514.00	8.0000	533.91	28.740	13.776	94	-0.69	028.42	0
028	Manganese (mg / kg (ppm))	0175	515.00	10.000	533.91	28.740	13.776	94	-0.66	028.31	0
028	Manganese (mg / kg (ppm))	0003	515.00	18.000	533.91	28.740	13.776	94	-0.66	028.41	0
028	Manganese (mg / kg (ppm))	0353	515.00	4.0000	533.91	28.740	13.776	94	-0.66	028.43	0
028	Manganese (mg / kg (ppm))	2128	516.41	6.9070	533.91	28.740	13.776	94	-0.61	028.32	0
028	Manganese (mg / kg (ppm))	2145	518.46	4.0200	533.91	28.740	13.776	94	-0.54	028.41	0
028	Manganese (mg / kg (ppm))	0870	519.25	9.9000	533.91	28.740	13.776	94	-0.51	028.42	0
028	Manganese (mg / kg (ppm))	0074	520.00	6.0000	533.91	28.740	13.776	94	-0.48	028.41	0
028	Manganese (mg / kg (ppm))	0689	521.10	31.800	533.91	28.740	13.776	94	-0.45	028.31	0
028	Manganese (mg / kg (ppm))	0019	521.10	16.200	533.91	28.740	13.776	94	-0.45	028.41	0
028	Manganese (mg / kg (ppm))	2022	522.00	2.6000	533.91	28.740	13.776	94	-0.41	028.31	0
028	Manganese (mg / kg (ppm))	0722	523.86	6.6497	533.91	28.740	13.776	94	-0.35	028.31	0
028	Manganese (mg / kg (ppm))	0948	524.95	1.9430	533.91	28.740	13.776	94	-0.31	028.33	0
028	Manganese (mg / kg (ppm))	0878	526.00	18.000	533.91	28.740	13.776	94	-0.28	028.41	0
028	Manganese (mg / kg (ppm))	0208	529.30	29.800	533.91	28.740	13.776	94	-0.16	028.41	0
028	Manganese (mg / kg (ppm))	0169	530.00	2.0000	533.91	28.740	13.776	94	-0.14	028.43	0
028	Manganese (mg / kg (ppm))	2053	533.00	34.000	533.91	28.740	13.776	94	-0.03	028.42	0
028	Manganese (mg / kg (ppm))	0148	533.05	22.750	533.91	28.740	13.776	94	-0.03	028.41	0
028	Manganese (mg / kg (ppm))	2097	534.41	0.85660	533.91	28.740	13.776	94	0.02	028.34	0
028	Manganese (mg / kg (ppm))	0668	534.50	33.000	533.91	28.740	13.776	94	0.02	028.43	0
028	Manganese (mg / kg (ppm))	0038	534.88	2.1900	533.91	28.740	13.776	94	0.03	028.43	0
028	Manganese (mg / kg (ppm))	0218	535.00	30.000	533.91	28.740	13.776	94	0.04	028.43	0
028	Manganese (mg / kg (ppm))	0190	535.44	6.3600	533.91	28.740	13.776	94	0.05	028.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
028	Manganese (mg / kg (ppm))	0098	536.35	13.500	533.91	28.740	13.776	94	0.08	028.43	0
028	Manganese (mg / kg (ppm))	0009	536.50	49.000	533.91	28.740	13.776	94	0.09	028.42	0
028	Manganese (mg / kg (ppm))	0675	537.41	2.5800	533.91	28.740	13.776	94	0.12	028.43	0
028	Manganese (mg / kg (ppm))	0693	537.50	3.4920	533.91	28.740	13.776	94	0.12	028.42	0
028	Manganese (mg / kg (ppm))	0413	538.00	6.0000	533.91	28.740	13.776	94	0.14	028.42	0
028	Manganese (mg / kg (ppm))	0868	538.51	0.14000	533.91	28.740	13.776	94	0.16	028.31	0
028	Manganese (mg / kg (ppm))	0042	541.00	30.000	533.91	28.740	13.776	94	0.25	028.42	0
028	Manganese (mg / kg (ppm))	0186	542.00	6.0000	533.91	28.740	13.776	94	0.28	028.42	0
028	Manganese (mg / kg (ppm))	0425	542.00	8.0000	533.91	28.740	13.776	94	0.28	028.43	0
028	Manganese (mg / kg (ppm))	0187	542.21	1.7900	533.91	28.740	13.776	94	0.29	028.42	0
028	Manganese (mg / kg (ppm))	0033	542.50	27.000	533.91	28.740	13.776	94	0.30	028.43	0
028	Manganese (mg / kg (ppm))	0598	542.75	12.700	533.91	28.740	13.776	94	0.31	028.41	0
028	Manganese (mg / kg (ppm))	0560	543.10	1.8000	533.91	28.740	13.776	94	0.32	028.42	0
028	Manganese (mg / kg (ppm))	2196	544.30	0.00000	533.91	28.740	13.776	94	0.36	028.31	0
028	Manganese (mg / kg (ppm))	0035	545.00	8.0000	533.91	28.740	13.776	94	0.39	028.42	0
028	Manganese (mg / kg (ppm))	0229	545.45	18.420	533.91	28.740	13.776	94	0.40	028.41	0
028	Manganese (mg / kg (ppm))	0242	545.50	15.000	533.91	28.740	13.776	94	0.40	028.99	0
028	Manganese (mg / kg (ppm))	0047	545.84	8.1500	533.91	28.740	13.776	94	0.42	028.52	0
028	Manganese (mg / kg (ppm))	0964	545.86	65.443	533.91	28.740	13.776	94	0.42	028.43	0
028	Manganese (mg / kg (ppm))	0939	546.09	18.600	533.91	28.740	13.776	94	0.42	028.31	0
028	Manganese (mg / kg (ppm))	0682	546.15	0.00000	533.91	28.740	13.776	94	0.43	028.43	0
028	Manganese (mg / kg (ppm))	0510	547.00	0.00000	533.91	28.740	13.776	94	0.46	028.43	0
028	Manganese (mg / kg (ppm))	0083	547.50	9.0000	533.91	28.740	13.776	94	0.47	028.43	0
028	Manganese (mg / kg (ppm))	0553	547.50	15.000	533.91	28.740	13.776	94	0.47	028.53	0
028	Manganese (mg / kg (ppm))	2188	549.54	5.5300	533.91	28.740	13.776	94	0.54	028.31	0
028	Manganese (mg / kg (ppm))	0505	551.20	7.0000	533.91	28.740	13.776	94	0.60	028.43	0
028	Manganese (mg / kg (ppm))	0358	554.41	36.780	533.91	28.740	13.776	94	0.71	028.41	0
028	Manganese (mg / kg (ppm))	0013	554.50	3.0000	533.91	28.740	13.776	94	0.72	028.43	0
028	Manganese (mg / kg (ppm))	0186	554.50	29.000	533.91	28.740	13.776	94	0.72	028.52	0
028	Manganese (mg / kg (ppm))	0572	554.50	5.0000	533.91	28.740	13.776	94	0.72	028.53	0
028	Manganese (mg / kg (ppm))	0354	555.46	5.3800	533.91	28.740	13.776	94	0.75	028.31	0
028	Manganese (mg / kg (ppm))	0910	556.50	13.000	533.91	28.740	13.776	94	0.79	028.41	0
028	Manganese (mg / kg (ppm))	0035	557.00	24.000	533.91	28.740	13.776	94	0.80	028.43	0
028	Manganese (mg / kg (ppm))	0034	559.25	32.300	533.91	28.740	13.776	94	0.88	028.43	0
028	Manganese (mg / kg (ppm))	0038	560.00	46.000	533.91	28.740	13.776	94	0.91	028.32	0
028	Manganese (mg / kg (ppm))	0202	561.49	10.700	533.91	28.740	13.776	94	0.96	028.43	0
028	Manganese (mg / kg (ppm))	0357	562.22	7.8270	533.91	28.740	13.776	94	0.98	028.42	0
028	Manganese (mg / kg (ppm))	0100	563.50	23.000	533.91	28.740	13.776	94	1.03	028.99	0
028	Manganese (mg / kg (ppm))	0001	566.75	6.5000	533.91	28.740	13.776	94	1.14	028.31	0
028	Manganese (mg / kg (ppm))	0208	569.00	28.000	533.91	28.740	13.776	94	1.22	028.31	0
028	Manganese (mg / kg (ppm))	0918	569.71	9.2600	533.91	28.740	13.776	94	1.25	028.43	0
028	Manganese (mg / kg (ppm))	0032	577.10	3.8000	533.91	28.740	13.776	94	1.50	028.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
028	Manganese (mg / kg (ppm))	0027	579.19	44.224	533.91	28.740	13.776	94	1.58	028.43	0
028	Manganese (mg / kg (ppm))	0294	584.96	4.1200	533.91	28.740	13.776	94	1.78	028.42	0
028	Manganese (mg / kg (ppm))	0017	585.12	4.7020	533.91	28.740	13.776	94	1.78	028.43	0
028	Manganese (mg / kg (ppm))	0405	596.00	0.00000	533.91	28.740	13.776	94	2.16	028.41	0
028	Manganese (mg / kg (ppm))	2144	602.71	2.4100	533.91	28.740	13.776	94	2.39	028.99	0
028	Manganese (mg / kg (ppm))	0610	605.50	33.000	533.91	28.740	13.776	94	2.49	028.43	0
028	Manganese (mg / kg (ppm))	2089	605.53	3.0000	533.91	28.740	13.776	94	2.49	028.43	0
028	Manganese (mg / kg (ppm))	0045	531.50	67.000	533.91	28.740	13.776	94	-0.08	028.42	1
031	Phosphorus (%)	0675	0.60500	0.01000	1.0871	0.05253	0.02979	137	-9.18	031.43	0
031	Phosphorus (%)	0108	0.85500	0.03000	1.0871	0.05253	0.02979	137	-4.42	031.01	0
031	Phosphorus (%)	0921	0.91095	0.07750	1.0871	0.05253	0.02979	137	-3.35	031.01	0
031	Phosphorus (%)	2055	0.92115	0.07550	1.0871	0.05253	0.02979	137	-3.16	031.53	0
031	Phosphorus (%)	0511	0.93500	0.01000	1.0871	0.05253	0.02979	137	-2.90	031.01	0
031	Phosphorus (%)	0278	0.95000	0.02000	1.0871	0.05253	0.02979	137	-2.61	031.42	0
031	Phosphorus (%)	0889	0.95400	0.03160	1.0871	0.05253	0.02979	137	-2.53	031.99	0
031	Phosphorus (%)	0186	0.95950	0.00500	1.0871	0.05253	0.02979	137	-2.43	031.52	0
031	Phosphorus (%)	0345	0.96800	0.02600	1.0871	0.05253	0.02979	137	-2.27	031.43	0
031	Phosphorus (%)	0358	0.97500	0.05000	1.0871	0.05253	0.02979	137	-2.13	031.41	0
031	Phosphorus (%)	0852	0.98000	0.04000	1.0871	0.05253	0.02979	137	-2.04	031.99	0
031	Phosphorus (%)	0619	0.98850	0.02300	1.0871	0.05253	0.02979	137	-1.88	031.41	0
031	Phosphorus (%)	0874	0.98900	0.00780	1.0871	0.05253	0.02979	137	-1.87	031.01	0
031	Phosphorus (%)	2200	1.0000	0.00000	1.0871	0.05253	0.02979	137	-1.66	031.01	0
031	Phosphorus (%)	2193	1.0050	0.01000	1.0871	0.05253	0.02979	137	-1.56	031.01	0
031	Phosphorus (%)	0674	1.0100	0.00000	1.0871	0.05253	0.02979	137	-1.47	031.01	0
031	Phosphorus (%)	0910	1.0100	0.04000	1.0871	0.05253	0.02979	137	-1.47	031.41	0
031	Phosphorus (%)	0596	1.0150	0.07000	1.0871	0.05253	0.02979	137	-1.37	031.01	0
031	Phosphorus (%)	0186	1.0150	0.01000	1.0871	0.05253	0.02979	137	-1.37	031.42	0
031	Phosphorus (%)	0037	1.0200	0.16000	1.0871	0.05253	0.02979	137	-1.28	031.42	0
031	Phosphorus (%)	0027	1.0240	0.07200	1.0871	0.05253	0.02979	137	-1.20	031.43	0
031	Phosphorus (%)	2031	1.0268	0.02260	1.0871	0.05253	0.02979	137	-1.15	031.01	0
031	Phosphorus (%)	0300	1.0300	0.08000	1.0871	0.05253	0.02979	137	-1.09	031.53	0
031	Phosphorus (%)	0047	1.0350	0.03000	1.0871	0.05253	0.02979	137	-0.99	031.42	0
031	Phosphorus (%)	0294	1.0400	0.00000	1.0871	0.05253	0.02979	137	-0.90	031.42	0
031	Phosphorus (%)	0152	1.0450	0.05000	1.0871	0.05253	0.02979	137	-0.80	031.01	0
031	Phosphorus (%)	0265	1.0450	0.07000	1.0871	0.05253	0.02979	137	-0.80	031.42	0
031	Phosphorus (%)	2199	1.0500	0.10000	1.0871	0.05253	0.02979	137	-0.71	031.01	0
031	Phosphorus (%)	0142	1.0513	0.01690	1.0871	0.05253	0.02979	137	-0.68	031.01	0
031	Phosphorus (%)	0622	1.0514	0.00000	1.0871	0.05253	0.02979	137	-0.68	031.01	0
031	Phosphorus (%)	0918	1.0520	0.01800	1.0871	0.05253	0.02979	137	-0.67	031.43	0
031	Phosphorus (%)	0407	1.0524	0.00130	1.0871	0.05253	0.02979	137	-0.66	031.41	0
031	Phosphorus (%)	0895	1.0550	0.01000	1.0871	0.05253	0.02979	137	-0.61	031.00	0
031	Phosphorus (%)	0144	1.0550	0.01000	1.0871	0.05253	0.02979	137	-0.61	031.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	0610	1.0550	0.03000	1.0871	0.05253	0.02979	137	-0.61	031.43	0
031	Phosphorus (%)	0202	1.0595	0.07900	1.0871	0.05253	0.02979	137	-0.53	031.43	0
031	Phosphorus (%)	0011	1.0637	0.00410	1.0871	0.05253	0.02979	137	-0.45	031.41	0
031	Phosphorus (%)	0870	1.0640	0.00800	1.0871	0.05253	0.02979	137	-0.44	031.42	0
031	Phosphorus (%)	0590	1.0650	0.03000	1.0871	0.05253	0.02979	137	-0.42	031.01	0
031	Phosphorus (%)	0019	1.0650	0.01000	1.0871	0.05253	0.02979	137	-0.42	031.41	0
031	Phosphorus (%)	0520	1.0650	0.05000	1.0871	0.05253	0.02979	137	-0.42	031.41	0
031	Phosphorus (%)	0010	1.0650	0.03000	1.0871	0.05253	0.02979	137	-0.42	031.43	0
031	Phosphorus (%)	0693	1.0660	0.08800	1.0871	0.05253	0.02979	137	-0.40	031.42	0
031	Phosphorus (%)	0722	1.0691	0.03990	1.0871	0.05253	0.02979	137	-0.34	031.01	0
031	Phosphorus (%)	0405	1.0700	0.00000	1.0871	0.05253	0.02979	137	-0.33	031.41	0
031	Phosphorus (%)	0685	1.0700	0.00000	1.0871	0.05253	0.02979	137	-0.33	031.41	0
031	Phosphorus (%)	2145	1.0742	0.01360	1.0871	0.05253	0.02979	137	-0.25	031.41	0
031	Phosphorus (%)	0233	1.0750	0.01000	1.0871	0.05253	0.02979	137	-0.23	031.01	0
031	Phosphorus (%)	0148	1.0750	0.01000	1.0871	0.05253	0.02979	137	-0.23	031.41	0
031	Phosphorus (%)	0100	1.0750	0.01000	1.0871	0.05253	0.02979	137	-0.23	031.99	0
031	Phosphorus (%)	0034	1.0795	0.05100	1.0871	0.05253	0.02979	137	-0.14	031.43	0
031	Phosphorus (%)	0968	1.0795	0.00300	1.0871	0.05253	0.02979	137	-0.14	031.43	0
031	Phosphorus (%)	0208	1.0800	0.00800	1.0871	0.05253	0.02979	137	-0.14	031.01	0
031	Phosphorus (%)	0939	1.0800	0.04000	1.0871	0.05253	0.02979	137	-0.14	031.01	0
031	Phosphorus (%)	2126	1.0800	0.02000	1.0871	0.05253	0.02979	137	-0.14	031.01	0
031	Phosphorus (%)	2192	1.0800	0.02000	1.0871	0.05253	0.02979	137	-0.14	031.01	0
031	Phosphorus (%)	0089	1.0800	0.00000	1.0871	0.05253	0.02979	137	-0.14	031.43	0
031	Phosphorus (%)	0682	1.0800	0.00000	1.0871	0.05253	0.02979	137	-0.14	031.43	0
031	Phosphorus (%)	2129	1.0805	0.02300	1.0871	0.05253	0.02979	137	-0.13	031.42	0
031	Phosphorus (%)	0651	1.0815	0.01500	1.0871	0.05253	0.02979	137	-0.11	031.01	0
031	Phosphorus (%)	0098	1.0815	0.04500	1.0871	0.05253	0.02979	137	-0.11	031.41	0
031	Phosphorus (%)	2097	1.0840	0.00530	1.0871	0.05253	0.02979	137	-0.06	031.01	0
031	Phosphorus (%)	0035	1.0850	0.01000	1.0871	0.05253	0.02979	137	-0.04	031.42	0
031	Phosphorus (%)	0881	1.0865	0.00500	1.0871	0.05253	0.02979	137	-0.01	031.01	0
031	Phosphorus (%)	0687	1.0900	0.02000	1.0871	0.05253	0.02979	137	0.06	031.01	0
031	Phosphorus (%)	0298	1.0900	0.02000	1.0871	0.05253	0.02979	137	0.06	031.41	0
031	Phosphorus (%)	0413	1.0900	0.04000	1.0871	0.05253	0.02979	137	0.06	031.41	0
031	Phosphorus (%)	2012	1.0900	0.00000	1.0871	0.05253	0.02979	137	0.06	031.41	0
031	Phosphorus (%)	0187	1.0900	0.00000	1.0871	0.05253	0.02979	137	0.06	031.42	0
031	Phosphorus (%)	2144	1.0900	0.02000	1.0871	0.05253	0.02979	137	0.06	031.99	0
031	Phosphorus (%)	0357	1.0941	0.00450	1.0871	0.05253	0.02979	137	0.13	031.42	0
031	Phosphorus (%)	2188	1.0950	0.01000	1.0871	0.05253	0.02979	137	0.15	031.01	0
031	Phosphorus (%)	0164	1.0950	0.01000	1.0871	0.05253	0.02979	137	0.15	031.41	0
031	Phosphorus (%)	0208	1.0950	0.02600	1.0871	0.05253	0.02979	137	0.15	031.41	0
031	Phosphorus (%)	0017	1.0960	0.07800	1.0871	0.05253	0.02979	137	0.17	031.43	0
031	Phosphorus (%)	0849	1.0975	0.00500	1.0871	0.05253	0.02979	137	0.20	031.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	0038	1.0975	0.00500	1.0871	0.05253	0.02979	137	0.20	031.43	0
031	Phosphorus (%)	0036	1.0986	0.00110	1.0871	0.05253	0.02979	137	0.22	031.03	0
031	Phosphorus (%)	0354	1.1000	0.00000	1.0871	0.05253	0.02979	137	0.25	031.01	0
031	Phosphorus (%)	0878	1.1000	0.00000	1.0871	0.05253	0.02979	137	0.25	031.01	0
031	Phosphorus (%)	2196	1.1000	0.00000	1.0871	0.05253	0.02979	137	0.25	031.01	0
031	Phosphorus (%)	0123	1.1000	0.00000	1.0871	0.05253	0.02979	137	0.25	031.41	0
031	Phosphorus (%)	0190	1.1000	0.02000	1.0871	0.05253	0.02979	137	0.25	031.42	0
031	Phosphorus (%)	0218	1.1000	0.00000	1.0871	0.05253	0.02979	137	0.25	031.43	0
031	Phosphorus (%)	0560	1.1018	0.01690	1.0871	0.05253	0.02979	137	0.28	031.42	0
031	Phosphorus (%)	2190	1.1042	0.00940	1.0871	0.05253	0.02979	137	0.33	031.01	0
031	Phosphorus (%)	0653	1.1045	0.01700	1.0871	0.05253	0.02979	137	0.33	031.41	0
031	Phosphorus (%)	0171	1.1050	0.01000	1.0871	0.05253	0.02979	137	0.34	031.41	0
031	Phosphorus (%)	0033	1.1050	0.01000	1.0871	0.05253	0.02979	137	0.34	031.43	0
031	Phosphorus (%)	0425	1.1050	0.01000	1.0871	0.05253	0.02979	137	0.34	031.43	0
031	Phosphorus (%)	0861	1.1050	0.03000	1.0871	0.05253	0.02979	137	0.34	031.43	0
031	Phosphorus (%)	0242	1.1050	0.07000	1.0871	0.05253	0.02979	137	0.34	031.99	0
031	Phosphorus (%)	0036	1.1058	0.08040	1.0871	0.05253	0.02979	137	0.36	031.43	0
031	Phosphorus (%)	2009	1.1065	0.00020	1.0871	0.05253	0.02979	137	0.37	031.01	0
031	Phosphorus (%)	0001	1.1085	0.01100	1.0871	0.05253	0.02979	137	0.41	031.03	0
031	Phosphorus (%)	0169	1.1100	0.04000	1.0871	0.05253	0.02979	137	0.44	031.03	0
031	Phosphorus (%)	0689	1.1100	0.02000	1.0871	0.05253	0.02979	137	0.44	031.01	0
031	Phosphorus (%)	2053	1.1100	0.02000	1.0871	0.05253	0.02979	137	0.44	031.02	0
031	Phosphorus (%)	0229	1.1110	0.04200	1.0871	0.05253	0.02979	137	0.46	031.41	0
031	Phosphorus (%)	2006	1.1135	0.03100	1.0871	0.05253	0.02979	137	0.50	031.01	0
031	Phosphorus (%)	0650	1.1150	0.03000	1.0871	0.05253	0.02979	137	0.53	031.01	0
031	Phosphorus (%)	0035	1.1150	0.07000	1.0871	0.05253	0.02979	137	0.53	031.43	0
031	Phosphorus (%)	2195	1.1150	0.01000	1.0871	0.05253	0.02979	137	0.53	031.01	0
031	Phosphorus (%)	2109	1.1190	0.03800	1.0871	0.05253	0.02979	137	0.61	031.01	0
031	Phosphorus (%)	0009	1.1202	0.03220	1.0871	0.05253	0.02979	137	0.63	031.42	0
031	Phosphorus (%)	0139	1.1235	0.00700	1.0871	0.05253	0.02979	137	0.69	031.41	0
031	Phosphorus (%)	0868	1.1250	0.01000	1.0871	0.05253	0.02979	137	0.72	031.01	0
031	Phosphorus (%)	0897	1.1250	0.03000	1.0871	0.05253	0.02979	137	0.72	031.01	0
031	Phosphorus (%)	0049	1.1250	0.07000	1.0871	0.05253	0.02979	137	0.72	031.41	0
031	Phosphorus (%)	0205	1.1250	0.01000	1.0871	0.05253	0.02979	137	0.72	031.42	0
031	Phosphorus (%)	0948	1.1255	0.01100	1.0871	0.05253	0.02979	137	0.73	031.01	0
031	Phosphorus (%)	0626	1.1290	0.02800	1.0871	0.05253	0.02979	137	0.80	031.01	0
031	Phosphorus (%)	0536	1.1300	0.02000	1.0871	0.05253	0.02979	137	0.82	031.06	0
031	Phosphorus (%)	0014	1.1300	0.06000	1.0871	0.05253	0.02979	137	0.82	031.42	0
031	Phosphorus (%)	0668	1.1300	0.06000	1.0871	0.05253	0.02979	137	0.82	031.43	0
031	Phosphorus (%)	2146	1.1300	0.00000	1.0871	0.05253	0.02979	137	0.82	031.51	0
031	Phosphorus (%)	0042	1.1350	0.11000	1.0871	0.05253	0.02979	137	0.91	031.42	0
031	Phosphorus (%)	0964	1.1376	0.07520	1.0871	0.05253	0.02979	137	0.96	031.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	0083	1.1400	0.02000	1.0871	0.05253	0.02979	137	1.01	031.43	0
031	Phosphorus (%)	0038	1.1418	0.03550	1.0871	0.05253	0.02979	137	1.04	031.01	0
031	Phosphorus (%)	0553	1.1500	0.02000	1.0871	0.05253	0.02979	137	1.20	031.53	0
031	Phosphorus (%)	2128	1.1540	0.21400	1.0871	0.05253	0.02979	137	1.27	031.01	0
031	Phosphorus (%)	0003	1.1550	0.09000	1.0871	0.05253	0.02979	137	1.29	031.41	0
031	Phosphorus (%)	0695	1.1565	0.02500	1.0871	0.05253	0.02979	137	1.32	031.41	0
031	Phosphorus (%)	0074	1.1600	0.02000	1.0871	0.05253	0.02979	137	1.39	031.41	0
031	Phosphorus (%)	0045	1.1600	0.08000	1.0871	0.05253	0.02979	137	1.39	031.42	0
031	Phosphorus (%)	0154	1.1628	0.11780	1.0871	0.05253	0.02979	137	1.44	031.52	0
031	Phosphorus (%)	0505	1.1650	0.01000	1.0871	0.05253	0.02979	137	1.48	031.02	0
031	Phosphorus (%)	0226	1.1650	0.03000	1.0871	0.05253	0.02979	137	1.48	031.41	0
031	Phosphorus (%)	0353	1.1650	0.01000	1.0871	0.05253	0.02979	137	1.48	031.43	0
031	Phosphorus (%)	0572	1.1800	0.04000	1.0871	0.05253	0.02979	137	1.77	031.53	0
031	Phosphorus (%)	0098	1.1845	0.00900	1.0871	0.05253	0.02979	137	1.85	031.43	0
031	Phosphorus (%)	0598	1.1935	0.01100	1.0871	0.05253	0.02979	137	2.03	031.41	0
031	Phosphorus (%)	2022	1.2050	0.01000	1.0871	0.05253	0.02979	137	2.24	031.01	0
031	Phosphorus (%)	0032	1.2165	0.05500	1.0871	0.05253	0.02979	137	2.46	031.43	0
031	Phosphorus (%)	2089	1.2300	0.08000	1.0871	0.05253	0.02979	137	2.72	031.43	0
031	Phosphorus (%)	0510	1.4500	0.02000	1.0871	0.05253	0.02979	137	6.91	031.43	0
031	Phosphorus (%)	0175	55.535	108.93	1.0871	0.05253	0.02979	137	1036.55	031.01	2
032	Potassium (%)	0889	0.51275	0.04930	1.2999	0.05367	0.03616	100	-14.67	032.99	0
032	Potassium (%)	0169	1.0800	0.00000	1.2999	0.05367	0.03616	100	-4.10	032.32	0
032	Potassium (%)	0619	1.1300	0.02000	1.2999	0.05367	0.03616	100	-3.17	032.41	0
032	Potassium (%)	0918	1.1855	0.09100	1.2999	0.05367	0.03616	100	-2.13	032.43	0
032	Potassium (%)	0144	1.1950	0.05000	1.2999	0.05367	0.03616	100	-1.96	032.41	0
032	Potassium (%)	0186	1.1950	0.01000	1.2999	0.05367	0.03616	100	-1.96	032.52	0
032	Potassium (%)	0536	1.2100	0.04000	1.2999	0.05367	0.03616	100	-1.68	032.31	0
032	Potassium (%)	0148	1.2150	0.09000	1.2999	0.05367	0.03616	100	-1.58	032.41	0
032	Potassium (%)	0042	1.2200	0.08000	1.2999	0.05367	0.03616	100	-1.49	032.42	0
032	Potassium (%)	0011	1.2260	0.00000	1.2999	0.05367	0.03616	100	-1.38	032.41	0
032	Potassium (%)	0229	1.2285	0.03700	1.2999	0.05367	0.03616	100	-1.33	032.41	0
032	Potassium (%)	0910	1.2300	0.04000	1.2999	0.05367	0.03616	100	-1.30	032.41	0
032	Potassium (%)	0100	1.2300	0.04000	1.2999	0.05367	0.03616	100	-1.30	032.99	0
032	Potassium (%)	0154	1.2310	0.29220	1.2999	0.05367	0.03616	100	-1.28	032.52	0
032	Potassium (%)	2055	1.2321	0.11180	1.2999	0.05367	0.03616	100	-1.26	032.53	0
032	Potassium (%)	0142	1.2395	0.02900	1.2999	0.05367	0.03616	100	-1.13	032.31	0
032	Potassium (%)	0868	1.2400	0.00000	1.2999	0.05367	0.03616	100	-1.12	032.31	0
032	Potassium (%)	0294	1.2550	0.01000	1.2999	0.05367	0.03616	100	-0.84	032.42	0
032	Potassium (%)	0175	1.2600	0.04000	1.2999	0.05367	0.03616	100	-0.74	032.31	0
032	Potassium (%)	2012	1.2600	0.00000	1.2999	0.05367	0.03616	100	-0.74	032.41	0
032	Potassium (%)	0405	1.2650	0.00000	1.2999	0.05367	0.03616	100	-0.65	032.41	0
032	Potassium (%)	0037	1.2650	0.23000	1.2999	0.05367	0.03616	100	-0.65	032.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
032	Potassium (%)	0047	1.2650	0.01000	1.2999	0.05367	0.03616	100	-0.65	032.42	0
032	Potassium (%)	0505	1.2650	0.01000	1.2999	0.05367	0.03616	100	-0.65	032.43	0
032	Potassium (%)	0964	1.2653	0.05980	1.2999	0.05367	0.03616	100	-0.65	032.43	0
032	Potassium (%)	0693	1.2700	0.03200	1.2999	0.05367	0.03616	100	-0.56	032.42	0
032	Potassium (%)	0013	1.2700	0.00000	1.2999	0.05367	0.03616	100	-0.56	032.43	0
032	Potassium (%)	0881	1.2705	0.00900	1.2999	0.05367	0.03616	100	-0.55	032.02	0
032	Potassium (%)	0948	1.2735	0.00500	1.2999	0.05367	0.03616	100	-0.49	032.31	0
032	Potassium (%)	0354	1.2750	0.01000	1.2999	0.05367	0.03616	100	-0.46	032.31	0
032	Potassium (%)	0049	1.2750	0.03000	1.2999	0.05367	0.03616	100	-0.46	032.41	0
032	Potassium (%)	0045	1.2750	0.15000	1.2999	0.05367	0.03616	100	-0.46	032.42	0
032	Potassium (%)	0139	1.2800	0.00600	1.2999	0.05367	0.03616	100	-0.37	032.41	0
032	Potassium (%)	0413	1.2800	0.00000	1.2999	0.05367	0.03616	100	-0.37	032.41	0
032	Potassium (%)	0653	1.2840	0.00800	1.2999	0.05367	0.03616	100	-0.30	032.41	0
032	Potassium (%)	0038	1.2850	0.02600	1.2999	0.05367	0.03616	100	-0.28	032.31	0
032	Potassium (%)	0186	1.2850	0.01000	1.2999	0.05367	0.03616	100	-0.28	032.42	0
032	Potassium (%)	2129	1.2870	0.06000	1.2999	0.05367	0.03616	100	-0.24	032.42	0
032	Potassium (%)	0032	1.2885	0.04300	1.2999	0.05367	0.03616	100	-0.21	032.43	0
032	Potassium (%)	0504	1.2900	0.02000	1.2999	0.05367	0.03616	100	-0.19	032.02	0
032	Potassium (%)	0003	1.2900	0.04000	1.2999	0.05367	0.03616	100	-0.19	032.41	0
032	Potassium (%)	0511	1.2900	0.08000	1.2999	0.05367	0.03616	100	-0.19	032.41	0
032	Potassium (%)	0878	1.2900	0.04000	1.2999	0.05367	0.03616	100	-0.19	032.41	0
032	Potassium (%)	0682	1.2900	0.00000	1.2999	0.05367	0.03616	100	-0.19	032.43	0
032	Potassium (%)	0861	1.2900	0.00000	1.2999	0.05367	0.03616	100	-0.19	032.43	0
032	Potassium (%)	2146	1.2900	0.04000	1.2999	0.05367	0.03616	100	-0.19	032.51	0
032	Potassium (%)	0553	1.2900	0.04000	1.2999	0.05367	0.03616	100	-0.19	032.53	0
032	Potassium (%)	0650	1.2950	0.01000	1.2999	0.05367	0.03616	100	-0.09	032.31	0
032	Potassium (%)	0035	1.2950	0.03000	1.2999	0.05367	0.03616	100	-0.09	032.42	0
032	Potassium (%)	0019	1.3000	0.02000	1.2999	0.05367	0.03616	100	0.00	032.41	0
032	Potassium (%)	0098	1.3000	0.02000	1.2999	0.05367	0.03616	100	0.00	032.41	0
032	Potassium (%)	0035	1.3000	0.04000	1.2999	0.05367	0.03616	100	0.00	032.43	0
032	Potassium (%)	0218	1.3000	0.00000	1.2999	0.05367	0.03616	100	0.00	032.43	0
032	Potassium (%)	0870	1.3025	0.00510	1.2999	0.05367	0.03616	100	0.05	032.42	0
032	Potassium (%)	0208	1.3030	0.00800	1.2999	0.05367	0.03616	100	0.06	032.41	0
032	Potassium (%)	0265	1.3050	0.07000	1.2999	0.05367	0.03616	100	0.09	032.42	0
032	Potassium (%)	0345	1.3050	0.03000	1.2999	0.05367	0.03616	100	0.09	032.43	0
032	Potassium (%)	0202	1.3055	0.00700	1.2999	0.05367	0.03616	100	0.10	032.43	0
032	Potassium (%)	0098	1.3100	0.10000	1.2999	0.05367	0.03616	100	0.19	032.43	0
032	Potassium (%)	2053	1.3150	0.01000	1.2999	0.05367	0.03616	100	0.28	032.31	0
032	Potassium (%)	0590	1.3150	0.01000	1.2999	0.05367	0.03616	100	0.28	032.99	0
032	Potassium (%)	0010	1.3200	0.10000	1.2999	0.05367	0.03616	100	0.37	032.43	0
032	Potassium (%)	0033	1.3200	0.04000	1.2999	0.05367	0.03616	100	0.37	032.43	0
032	Potassium (%)	0034	1.3240	0.05200	1.2999	0.05367	0.03616	100	0.45	032.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
032	Potassium (%)	0353	1.3250	0.01000	1.2999	0.05367	0.03616	100	0.47	032.43	0
032	Potassium (%)	0027	1.3255	0.11900	1.2999	0.05367	0.03616	100	0.48	032.43	0
032	Potassium (%)	0407	1.3279	0.01690	1.2999	0.05367	0.03616	100	0.52	032.41	0
032	Potassium (%)	0208	1.3300	0.02000	1.2999	0.05367	0.03616	100	0.56	032.31	0
032	Potassium (%)	0278	1.3300	0.00000	1.2999	0.05367	0.03616	100	0.56	032.42	0
032	Potassium (%)	0009	1.3301	0.07390	1.2999	0.05367	0.03616	100	0.56	032.42	0
032	Potassium (%)	0164	1.3350	0.07000	1.2999	0.05367	0.03616	100	0.65	032.41	0
032	Potassium (%)	0038	1.3350	0.03000	1.2999	0.05367	0.03616	100	0.65	032.43	0
032	Potassium (%)	0510	1.3350	0.01000	1.2999	0.05367	0.03616	100	0.65	032.43	0
032	Potassium (%)	0074	1.3400	0.00000	1.2999	0.05367	0.03616	100	0.75	032.41	0
032	Potassium (%)	0358	1.3400	0.02000	1.2999	0.05367	0.03616	100	0.75	032.41	0
032	Potassium (%)	0520	1.3400	0.00000	1.2999	0.05367	0.03616	100	0.75	032.41	0
032	Potassium (%)	0187	1.3400	0.00000	1.2999	0.05367	0.03616	100	0.75	032.42	0
032	Potassium (%)	0190	1.3400	0.02000	1.2999	0.05367	0.03616	100	0.75	032.42	0
032	Potassium (%)	0560	1.3403	0.10740	1.2999	0.05367	0.03616	100	0.75	032.42	0
032	Potassium (%)	0357	1.3429	0.00270	1.2999	0.05367	0.03616	100	0.80	032.42	0
032	Potassium (%)	0695	1.3450	0.02800	1.2999	0.05367	0.03616	100	0.84	032.41	0
032	Potassium (%)	0083	1.3450	0.01000	1.2999	0.05367	0.03616	100	0.84	032.43	0
032	Potassium (%)	0425	1.3450	0.01000	1.2999	0.05367	0.03616	100	0.84	032.43	0
032	Potassium (%)	0668	1.3450	0.05000	1.2999	0.05367	0.03616	100	0.84	032.43	0
032	Potassium (%)	0675	1.3450	0.03000	1.2999	0.05367	0.03616	100	0.84	032.43	0
032	Potassium (%)	0001	1.3480	0.01200	1.2999	0.05367	0.03616	100	0.90	032.99	0
032	Potassium (%)	0939	1.3550	0.05000	1.2999	0.05367	0.03616	100	1.03	032.31	0
032	Potassium (%)	0205	1.3550	0.03000	1.2999	0.05367	0.03616	100	1.03	032.42	0
032	Potassium (%)	0610	1.3550	0.03000	1.2999	0.05367	0.03616	100	1.03	032.43	0
032	Potassium (%)	0171	1.3600	0.02000	1.2999	0.05367	0.03616	100	1.12	032.41	0
032	Potassium (%)	0598	1.3600	0.02800	1.2999	0.05367	0.03616	100	1.12	032.41	0
032	Potassium (%)	0612	1.3700	0.02000	1.2999	0.05367	0.03616	100	1.31	032.33	0
032	Potassium (%)	0242	1.3850	0.01000	1.2999	0.05367	0.03616	100	1.58	032.99	0
032	Potassium (%)	0968	1.4005	0.01500	1.2999	0.05367	0.03616	100	1.87	032.43	0
032	Potassium (%)	0226	1.4050	0.03000	1.2999	0.05367	0.03616	100	1.96	032.41	0
032	Potassium (%)	2089	1.4200	0.10000	1.2999	0.05367	0.03616	100	2.24	032.43	0
032	Potassium (%)	0152	1.4350	0.01000	1.2999	0.05367	0.03616	100	2.52	032.44	0
032	Potassium (%)	0572	1.4350	0.05000	1.2999	0.05367	0.03616	100	2.52	032.53	0
032	Potassium (%)	0017	1.4760	0.04200	1.2999	0.05367	0.03616	100	3.28	032.43	0
032	Potassium (%)	0089	1.5500	0.00000	1.2999	0.05367	0.03616	100	4.66	032.43	0
032	Potassium (%)	0874	64.645	126.71	1.2999	0.05367	0.03616	100	1180.16	032.31	2
033	Salt as chloride (%)	2145	1.9350	0.03000	3.4956	0.14131	0.07029	72	-11.04	033.00	0
033	Salt as chloride (%)	0298	2.1300	0.02000	3.4956	0.14131	0.07029	72	-9.66	033.00	0
033	Salt as chloride (%)	0171	2.7700	0.04000	3.4956	0.14131	0.07029	72	-5.14	033.00	0
033	Salt as chloride (%)	2193	2.9950	0.09000	3.4956	0.14131	0.07029	72	-3.54	033.00	0
033	Salt as chloride (%)	0358	3.2400	0.24000	3.4956	0.14131	0.07029	72	-1.81	033.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
033	Salt as chloride (%)	0265	3.3000	0.00000	3.4956	0.14131	0.07029	72	-1.38	033.03	0
033	Salt as chloride (%)	0504	3.3100	0.04000	3.4956	0.14131	0.07029	72	-1.31	033.00	0
033	Salt as chloride (%)	0675	3.3250	0.21000	3.4956	0.14131	0.07029	72	-1.21	033.00	0
033	Salt as chloride (%)	0511	3.3300	0.00000	3.4956	0.14131	0.07029	72	-1.17	033.00	0
033	Salt as chloride (%)	2196	3.3300	0.00000	3.4956	0.14131	0.07029	72	-1.17	033.00	0
033	Salt as chloride (%)	0881	3.3395	0.09900	3.4956	0.14131	0.07029	72	-1.10	033.00	0
033	Salt as chloride (%)	0353	3.3450	0.03000	3.4956	0.14131	0.07029	72	-1.07	033.00	0
033	Salt as chloride (%)	0619	3.3500	0.04000	3.4956	0.14131	0.07029	72	-1.03	033.01	0
033	Salt as chloride (%)	0539	3.3600	0.06000	3.4956	0.14131	0.07029	72	-0.96	033.00	0
033	Salt as chloride (%)	0229	3.3750	0.03000	3.4956	0.14131	0.07029	72	-0.85	033.01	0
033	Salt as chloride (%)	0027	3.3835	0.39300	3.4956	0.14131	0.07029	72	-0.79	033.99	0
033	Salt as chloride (%)	2109	3.4025	0.17300	3.4956	0.14131	0.07029	72	-0.66	033.01	0
033	Salt as chloride (%)	0169	3.4100	0.02000	3.4956	0.14131	0.07029	72	-0.61	033.00	0
033	Salt as chloride (%)	0164	3.4100	0.00000	3.4956	0.14131	0.07029	72	-0.61	033.01	0
033	Salt as chloride (%)	0045	3.4200	0.08000	3.4956	0.14131	0.07029	72	-0.54	033.00	0
033	Salt as chloride (%)	0921	3.4200	0.22000	3.4956	0.14131	0.07029	72	-0.54	033.00	0
033	Salt as chloride (%)	0190	3.4200	0.00000	3.4956	0.14131	0.07029	72	-0.54	033.03	0
033	Salt as chloride (%)	0100	3.4250	0.05000	3.4956	0.14131	0.07029	72	-0.50	033.01	0
033	Salt as chloride (%)	0011	3.4313	0.00000	3.4956	0.14131	0.07029	72	-0.46	033.01	0
033	Salt as chloride (%)	2009	3.4316	0.00570	3.4956	0.14131	0.07029	72	-0.45	033.00	0
033	Salt as chloride (%)	0878	3.4350	0.03000	3.4956	0.14131	0.07029	72	-0.43	033.00	0
033	Salt as chloride (%)	0003	3.4500	0.10000	3.4956	0.14131	0.07029	72	-0.32	033.01	0
033	Salt as chloride (%)	0868	3.4500	0.04000	3.4956	0.14131	0.07029	72	-0.32	033.05	0
033	Salt as chloride (%)	0910	3.4550	0.33000	3.4956	0.14131	0.07029	72	-0.29	033.99	0
033	Salt as chloride (%)	0098	3.4600	0.02000	3.4956	0.14131	0.07029	72	-0.25	033.01	0
033	Salt as chloride (%)	0897	3.4650	0.05000	3.4956	0.14131	0.07029	72	-0.22	033.00	0
033	Salt as chloride (%)	0175	3.4650	0.03000	3.4956	0.14131	0.07029	72	-0.22	033.01	0
033	Salt as chloride (%)	0590	3.4850	0.17000	3.4956	0.14131	0.07029	72	-0.08	033.01	0
033	Salt as chloride (%)	0354	3.4900	0.00000	3.4956	0.14131	0.07029	72	-0.04	033.01	0
033	Salt as chloride (%)	0413	3.5000	0.06000	3.4956	0.14131	0.07029	72	0.03	033.01	0
033	Salt as chloride (%)	0874	3.5050	0.01000	3.4956	0.14131	0.07029	72	0.07	033.01	0
033	Salt as chloride (%)	0939	3.5050	0.11000	3.4956	0.14131	0.07029	72	0.07	033.01	0
033	Salt as chloride (%)	0682	3.5100	0.00000	3.4956	0.14131	0.07029	72	0.10	033.99	0
033	Salt as chloride (%)	0693	3.5115	0.08700	3.4956	0.14131	0.07029	72	0.11	033.00	0
033	Salt as chloride (%)	0948	3.5150	0.01000	3.4956	0.14131	0.07029	72	0.14	033.01	0
033	Salt as chloride (%)	0689	3.5200	0.06000	3.4956	0.14131	0.07029	72	0.17	033.05	0
033	Salt as chloride (%)	0148	3.5400	0.00000	3.4956	0.14131	0.07029	72	0.31	033.01	0
033	Salt as chloride (%)	0010	3.5450	0.01000	3.4956	0.14131	0.07029	72	0.35	033.01	0
033	Salt as chloride (%)	0242	3.5450	0.03000	3.4956	0.14131	0.07029	72	0.35	033.01	0
033	Salt as chloride (%)	2022	3.5500	0.02000	3.4956	0.14131	0.07029	72	0.38	033.00	0
033	Salt as chloride (%)	0650	3.5600	0.02000	3.4956	0.14131	0.07029	72	0.46	033.01	0
033	Salt as chloride (%)	2181	3.5600	0.02000	3.4956	0.14131	0.07029	72	0.46	033.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
033	Salt as chloride (%)	0226	3.5600	0.00000	3.4956	0.14131	0.07029	72	0.46	033.01	0
033	Salt as chloride (%)	0861	3.5650	0.01000	3.4956	0.14131	0.07029	72	0.49	033.01	0
033	Salt as chloride (%)	2129	3.5650	0.09000	3.4956	0.14131	0.07029	72	0.49	033.05	0
033	Salt as chloride (%)	0004	3.5700	0.10000	3.4956	0.14131	0.07029	72	0.53	033.01	0
033	Salt as chloride (%)	0510	3.5700	0.02000	3.4956	0.14131	0.07029	72	0.53	033.01	0
033	Salt as chloride (%)	2190	3.5774	0.09900	3.4956	0.14131	0.07029	72	0.58	033.00	0
033	Salt as chloride (%)	0309	3.5775	0.05210	3.4956	0.14131	0.07029	72	0.58	033.00	0
033	Salt as chloride (%)	0407	3.5800	0.00000	3.4956	0.14131	0.07029	72	0.60	033.01	0
033	Salt as chloride (%)	0083	3.5850	0.07000	3.4956	0.14131	0.07029	72	0.63	033.01	0
033	Salt as chloride (%)	0001	3.5900	0.04000	3.4956	0.14131	0.07029	72	0.67	033.01	0
033	Salt as chloride (%)	2146	3.5900	0.00000	3.4956	0.14131	0.07029	72	0.67	033.01	0
033	Salt as chloride (%)	0895	3.6050	0.01000	3.4956	0.14131	0.07029	72	0.77	033.00	0
033	Salt as chloride (%)	0644	3.6175	0.05420	3.4956	0.14131	0.07029	72	0.86	033.01	0
033	Salt as chloride (%)	0205	3.6300	0.10000	3.4956	0.14131	0.07029	72	0.95	033.01	0
033	Salt as chloride (%)	0278	3.6400	0.04000	3.4956	0.14131	0.07029	72	1.02	033.01	0
033	Salt as chloride (%)	0028	3.6450	0.21000	3.4956	0.14131	0.07029	72	1.06	033.01	0
033	Salt as chloride (%)	0596	3.6600	0.00000	3.4956	0.14131	0.07029	72	1.16	033.01	0
033	Salt as chloride (%)	0425	3.6850	0.07000	3.4956	0.14131	0.07029	72	1.34	033.01	0
033	Salt as chloride (%)	0019	3.6900	0.22000	3.4956	0.14131	0.07029	72	1.38	033.01	0
033	Salt as chloride (%)	0017	3.6980	0.21800	3.4956	0.14131	0.07029	72	1.43	033.99	0
033	Salt as chloride (%)	0505	3.7000	0.02000	3.4956	0.14131	0.07029	72	1.45	033.03	0
033	Salt as chloride (%)	2144	3.7050	0.01000	3.4956	0.14131	0.07029	72	1.48	033.99	0
033	Salt as chloride (%)	2073	3.7150	0.15000	3.4956	0.14131	0.07029	72	1.55	033.99	0
033	Salt as chloride (%)	0964	3.7270	0.09000	3.4956	0.14131	0.07029	72	1.64	033.01	0
033	Salt as chloride (%)	0144	3.7450	0.31000	3.4956	0.14131	0.07029	72	1.76	033.03	0
033	Salt as chloride (%)	0853	4.1800	1.0200	3.4956	0.14131	0.07029	72	4.84	033.99	1
034	Selenium (mg / kg (ppm))	0011	2.8710	0.18000	5.6090	0.71442	0.19818	28	-3.83	034.41	0
034	Selenium (mg / kg (ppm))	0032	4.1500	0.20000	5.6090	0.71442	0.19818	28	-2.04	034.53	0
034	Selenium (mg / kg (ppm))	0619	4.1750	0.05000	5.6090	0.71442	0.19818	28	-2.01	034.41	0
034	Selenium (mg / kg (ppm))	2055	4.5750	0.17000	5.6090	0.71442	0.19818	28	-1.45	034.53	0
034	Selenium (mg / kg (ppm))	0190	4.6850	0.11000	5.6090	0.71442	0.19818	28	-1.29	034.99	0
034	Selenium (mg / kg (ppm))	0164	4.9500	0.02000	5.6090	0.71442	0.19818	28	-0.92	034.04	0
034	Selenium (mg / kg (ppm))	0910	5.3400	0.34000	5.6090	0.71442	0.19818	28	-0.38	034.52	0
034	Selenium (mg / kg (ppm))	0047	5.4200	0.20000	5.6090	0.71442	0.19818	28	-0.26	034.52	0
034	Selenium (mg / kg (ppm))	0407	5.4850	0.45000	5.6090	0.71442	0.19818	28	-0.17	034.43	0
034	Selenium (mg / kg (ppm))	0035	5.4875	0.03700	5.6090	0.71442	0.19818	28	-0.17	034.04	0
034	Selenium (mg / kg (ppm))	0218	5.5000	0.20000	5.6090	0.71442	0.19818	28	-0.15	034.53	0
034	Selenium (mg / kg (ppm))	0171	5.5649	0.27020	5.6090	0.71442	0.19818	28	-0.06	034.04	0
034	Selenium (mg / kg (ppm))	0169	5.5800	0.02000	5.6090	0.71442	0.19818	28	-0.04	034.04	0
034	Selenium (mg / kg (ppm))	0964	5.6161	0.44450	5.6090	0.71442	0.19818	28	0.01	034.43	0
034	Selenium (mg / kg (ppm))	0148	5.6400	0.00000	5.6090	0.71442	0.19818	28	0.04	034.41	0
034	Selenium (mg / kg (ppm))	0186	5.7050	0.25000	5.6090	0.71442	0.19818	28	0.13	034.52	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
034	Selenium (mg / kg (ppm))	0098	5.7670	0.51600	5.6090	0.71442	0.19818	28	0.22	034.53	0
034	Selenium (mg / kg (ppm))	0013	5.7700	0.02000	5.6090	0.71442	0.19818	28	0.23	034.53	0
034	Selenium (mg / kg (ppm))	0208	5.8500	0.12000	5.6090	0.71442	0.19818	28	0.34	034.52	0
034	Selenium (mg / kg (ppm))	0610	5.9400	0.22000	5.6090	0.71442	0.19818	28	0.46	034.04	0
034	Selenium (mg / kg (ppm))	0553	5.9450	0.33000	5.6090	0.71442	0.19818	28	0.47	034.53	0
034	Selenium (mg / kg (ppm))	0027	6.0495	0.00700	5.6090	0.71442	0.19818	28	0.62	034.43	0
034	Selenium (mg / kg (ppm))	0968	6.2350	0.05000	5.6090	0.71442	0.19818	28	0.88	034.43	0
034	Selenium (mg / kg (ppm))	0560	6.3050	0.17000	5.6090	0.71442	0.19818	28	0.97	034.53	0
034	Selenium (mg / kg (ppm))	0939	6.3100	0.02000	5.6090	0.71442	0.19818	28	0.98	034.31	0
034	Selenium (mg / kg (ppm))	0918	6.3579	0.26420	5.6090	0.71442	0.19818	28	1.05	034.53	0
034	Selenium (mg / kg (ppm))	0036	6.8500	0.30000	5.6090	0.71442	0.19818	28	1.74	034.43	0
034	Selenium (mg / kg (ppm))	2012	8.0350	0.59000	5.6090	0.71442	0.19818	28	3.40	034.53	0
034	Selenium (mg / kg (ppm))	0218	< 22		5.6090	0.71442	0.19818	28		034.43	5
035	Sodium (%)	0034	0.79095	0.03990	1.2609	0.07731	0.03859	98	-6.08	035.43	0
035	Sodium (%)	0169	1.0150	0.01000	1.2609	0.07731	0.03859	98	-3.18	035.32	0
035	Sodium (%)	0889	1.0461	0.06510	1.2609	0.07731	0.03859	98	-2.78	035.99	0
035	Sodium (%)	0098	1.0700	0.02000	1.2609	0.07731	0.03859	98	-2.47	035.41	0
035	Sodium (%)	0089	1.0700	0.00000	1.2609	0.07731	0.03859	98	-2.47	035.43	0
035	Sodium (%)	0175	1.1100	0.02000	1.2609	0.07731	0.03859	98	-1.95	035.31	0
035	Sodium (%)	0619	1.1350	0.03000	1.2609	0.07731	0.03859	98	-1.63	035.41	0
035	Sodium (%)	0298	1.1400	0.02000	1.2609	0.07731	0.03859	98	-1.56	035.41	0
035	Sodium (%)	0511	1.1400	0.16000	1.2609	0.07731	0.03859	98	-1.56	035.41	0
035	Sodium (%)	0148	1.1450	0.03000	1.2609	0.07731	0.03859	98	-1.50	035.41	0
035	Sodium (%)	0294	1.1500	0.02000	1.2609	0.07731	0.03859	98	-1.43	035.42	0
035	Sodium (%)	0042	1.1600	0.04000	1.2609	0.07731	0.03859	98	-1.31	035.42	0
035	Sodium (%)	0049	1.1650	0.15000	1.2609	0.07731	0.03859	98	-1.24	035.41	0
035	Sodium (%)	0142	1.1689	0.00670	1.2609	0.07731	0.03859	98	-1.19	035.31	0
035	Sodium (%)	2055	1.1700	0.11620	1.2609	0.07731	0.03859	98	-1.18	035.53	0
035	Sodium (%)	2012	1.1750	0.01000	1.2609	0.07731	0.03859	98	-1.11	035.41	0
035	Sodium (%)	2196	1.1800	0.00000	1.2609	0.07731	0.03859	98	-1.05	035.31	0
035	Sodium (%)	0536	1.1900	0.00000	1.2609	0.07731	0.03859	98	-0.92	035.31	0
035	Sodium (%)	0675	1.1950	0.05000	1.2609	0.07731	0.03859	98	-0.85	035.43	0
035	Sodium (%)	0139	1.1950	0.00400	1.2609	0.07731	0.03859	98	-0.85	035.41	0
035	Sodium (%)	0218	1.2000	0.00000	1.2609	0.07731	0.03859	98	-0.79	035.43	0
035	Sodium (%)	0590	1.2000	0.04000	1.2609	0.07731	0.03859	98	-0.79	035.99	0
035	Sodium (%)	0598	1.2075	0.07500	1.2609	0.07731	0.03859	98	-0.69	035.41	0
035	Sodium (%)	0354	1.2100	0.02000	1.2609	0.07731	0.03859	98	-0.66	035.31	0
035	Sodium (%)	0144	1.2100	0.08000	1.2609	0.07731	0.03859	98	-0.66	035.41	0
035	Sodium (%)	0164	1.2150	0.03000	1.2609	0.07731	0.03859	98	-0.59	035.41	0
035	Sodium (%)	0229	1.2160	0.04200	1.2609	0.07731	0.03859	98	-0.58	035.41	0
035	Sodium (%)	2195	1.2200	0.02000	1.2609	0.07731	0.03859	98	-0.53	035.31	0
035	Sodium (%)	0405	1.2200	0.00000	1.2609	0.07731	0.03859	98	-0.53	035.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
035	Sodium (%)	0186	1.2250	0.01000	1.2609	0.07731	0.03859	98	-0.46	035.42	0
035	Sodium (%)	0202	1.2325	0.02700	1.2609	0.07731	0.03859	98	-0.37	035.43	0
035	Sodium (%)	0345	1.2350	0.03000	1.2609	0.07731	0.03859	98	-0.33	035.43	0
035	Sodium (%)	0868	1.2400	0.00000	1.2609	0.07731	0.03859	98	-0.27	035.31	0
035	Sodium (%)	0171	1.2400	0.00000	1.2609	0.07731	0.03859	98	-0.27	035.41	0
035	Sodium (%)	0098	1.2400	0.04000	1.2609	0.07731	0.03859	98	-0.27	035.43	0
035	Sodium (%)	0100	1.2400	0.00000	1.2609	0.07731	0.03859	98	-0.27	035.99	0
035	Sodium (%)	0910	1.2450	0.03000	1.2609	0.07731	0.03859	98	-0.21	035.41	0
035	Sodium (%)	0881	1.2480	0.00400	1.2609	0.07731	0.03859	98	-0.17	035.01	0
035	Sodium (%)	0208	1.2500	0.02000	1.2609	0.07731	0.03859	98	-0.14	035.31	0
035	Sodium (%)	0650	1.2500	0.16000	1.2609	0.07731	0.03859	98	-0.14	035.31	0
035	Sodium (%)	2146	1.2500	0.02000	1.2609	0.07731	0.03859	98	-0.14	035.51	0
035	Sodium (%)	0868	1.2550	0.03000	1.2609	0.07731	0.03859	98	-0.08	035.01	0
035	Sodium (%)	0870	1.2550	0.04400	1.2609	0.07731	0.03859	98	-0.08	035.42	0
035	Sodium (%)	0653	1.2555	0.03700	1.2609	0.07731	0.03859	98	-0.07	035.41	0
035	Sodium (%)	2097	1.2599	0.03130	1.2609	0.07731	0.03859	98	-0.01	035.31	0
035	Sodium (%)	0019	1.2600	0.04000	1.2609	0.07731	0.03859	98	-0.01	035.41	0
035	Sodium (%)	0693	1.2635	0.06700	1.2609	0.07731	0.03859	98	0.03	035.42	0
035	Sodium (%)	0278	1.2650	0.03000	1.2609	0.07731	0.03859	98	0.05	035.42	0
035	Sodium (%)	0186	1.2650	0.01000	1.2609	0.07731	0.03859	98	0.05	035.52	0
035	Sodium (%)	0208	1.2655	0.00900	1.2609	0.07731	0.03859	98	0.06	035.41	0
035	Sodium (%)	0190	1.2700	0.06000	1.2609	0.07731	0.03859	98	0.12	035.42	0
035	Sodium (%)	0035	1.2750	0.01000	1.2609	0.07731	0.03859	98	0.18	035.42	0
035	Sodium (%)	2053	1.2750	0.05000	1.2609	0.07731	0.03859	98	0.18	035.42	0
035	Sodium (%)	0035	1.2750	0.07000	1.2609	0.07731	0.03859	98	0.18	035.43	0
035	Sodium (%)	0505	1.2750	0.01000	1.2609	0.07731	0.03859	98	0.18	035.43	0
035	Sodium (%)	0668	1.2750	0.01000	1.2609	0.07731	0.03859	98	0.18	035.43	0
035	Sodium (%)	0510	1.2760	0.01800	1.2609	0.07731	0.03859	98	0.20	035.43	0
035	Sodium (%)	2192	1.2800	0.04000	1.2609	0.07731	0.03859	98	0.25	035.43	0
035	Sodium (%)	0572	1.2800	0.00000	1.2609	0.07731	0.03859	98	0.25	035.53	0
035	Sodium (%)	0265	1.2900	0.16000	1.2609	0.07731	0.03859	98	0.38	035.42	0
035	Sodium (%)	0038	1.2910	0.02800	1.2609	0.07731	0.03859	98	0.39	035.31	0
035	Sodium (%)	0038	1.2925	0.03500	1.2609	0.07731	0.03859	98	0.41	035.43	0
035	Sodium (%)	0948	1.2930	0.01200	1.2609	0.07731	0.03859	98	0.42	035.31	0
035	Sodium (%)	0939	1.2950	0.05000	1.2609	0.07731	0.03859	98	0.44	035.31	0
035	Sodium (%)	0353	1.2950	0.05000	1.2609	0.07731	0.03859	98	0.44	035.43	0
035	Sodium (%)	0861	1.2950	0.07000	1.2609	0.07731	0.03859	98	0.44	035.43	0
035	Sodium (%)	0407	1.2963	0.04330	1.2609	0.07731	0.03859	98	0.46	035.41	0
035	Sodium (%)	0918	1.2985	0.08900	1.2609	0.07731	0.03859	98	0.49	035.43	0
035	Sodium (%)	0413	1.3000	0.00000	1.2609	0.07731	0.03859	98	0.51	035.41	0
035	Sodium (%)	0878	1.3000	0.06000	1.2609	0.07731	0.03859	98	0.51	035.41	0
035	Sodium (%)	0425	1.3000	0.04000	1.2609	0.07731	0.03859	98	0.51	035.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
035	Sodium (%)	0242	1.3000	0.02000	1.2609	0.07731	0.03859	98	0.51	035.99	0
035	Sodium (%)	0037	1.3050	0.07000	1.2609	0.07731	0.03859	98	0.57	035.05	0
035	Sodium (%)	0358	1.3050	0.03000	1.2609	0.07731	0.03859	98	0.57	035.41	0
035	Sodium (%)	0520	1.3079	0.12430	1.2609	0.07731	0.03859	98	0.61	035.41	0
035	Sodium (%)	0123	1.3100	0.00000	1.2609	0.07731	0.03859	98	0.64	035.41	0
035	Sodium (%)	0682	1.3100	0.00000	1.2609	0.07731	0.03859	98	0.64	035.43	0
035	Sodium (%)	0560	1.3116	0.08180	1.2609	0.07731	0.03859	98	0.66	035.42	0
035	Sodium (%)	0504	1.3150	0.05000	1.2609	0.07731	0.03859	98	0.70	035.05	0
035	Sodium (%)	0187	1.3200	0.00000	1.2609	0.07731	0.03859	98	0.76	035.42	0
035	Sodium (%)	2006	1.3235	0.01100	1.2609	0.07731	0.03859	98	0.81	035.01	0
035	Sodium (%)	0964	1.3243	0.07580	1.2609	0.07731	0.03859	98	0.82	035.43	0
035	Sodium (%)	0689	1.3300	0.02000	1.2609	0.07731	0.03859	98	0.89	035.31	0
035	Sodium (%)	0226	1.3300	0.01000	1.2609	0.07731	0.03859	98	0.89	035.41	0
035	Sodium (%)	0045	1.3300	0.06000	1.2609	0.07731	0.03859	98	0.89	035.42	0
035	Sodium (%)	0083	1.3350	0.03000	1.2609	0.07731	0.03859	98	0.96	035.43	0
035	Sodium (%)	0032	1.3400	0.02000	1.2609	0.07731	0.03859	98	1.02	035.43	0
035	Sodium (%)	2188	1.3450	0.01000	1.2609	0.07731	0.03859	98	1.09	035.31	0
035	Sodium (%)	0108	1.3700	0.10000	1.2609	0.07731	0.03859	98	1.41	035.05	0
035	Sodium (%)	2089	1.3800	0.12000	1.2609	0.07731	0.03859	98	1.54	035.43	0
035	Sodium (%)	0610	1.3900	0.02000	1.2609	0.07731	0.03859	98	1.67	035.43	0
035	Sodium (%)	0205	1.3950	0.03000	1.2609	0.07731	0.03859	98	1.73	035.42	0
035	Sodium (%)	0874	1.4000	0.06000	1.2609	0.07731	0.03859	98	1.80	035.31	0
035	Sodium (%)	0695	1.4045	0.00300	1.2609	0.07731	0.03859	98	1.86	035.41	0
035	Sodium (%)	0154	1.4095	0.03650	1.2609	0.07731	0.03859	98	1.92	035.52	0
035	Sodium (%)	2126	1.4200	0.02000	1.2609	0.07731	0.03859	98	2.06	035.31	0
035	Sodium (%)	0553	1.4250	0.05000	1.2609	0.07731	0.03859	98	2.12	035.53	0
035	Sodium (%)	0017	1.4550	0.08600	1.2609	0.07731	0.03859	98	2.51	035.43	0
035	Sodium (%)	0233	0.13000	0.00000	1.2609	0.07731	0.03859	98	-14.63	035.31	2
036	Sulfur (%)	0345	0.33500	0.01000	0.38150	0.02635	0.01223	38	-1.76	036.43	0
036	Sulfur (%)	0870	0.33565	0.00670	0.38150	0.02635	0.01223	38	-1.74	036.42	0
036	Sulfur (%)	0045	0.34050	0.02700	0.38150	0.02635	0.01223	38	-1.56	036.42	0
036	Sulfur (%)	0889	0.34495	0.00750	0.38150	0.02635	0.01223	38	-1.39	036.99	0
036	Sulfur (%)	0265	0.34500	0.03000	0.38150	0.02635	0.01223	38	-1.38	036.42	0
036	Sulfur (%)	0037	0.35500	0.03000	0.38150	0.02635	0.01223	38	-1.01	036.42	0
036	Sulfur (%)	0693	0.36350	0.00900	0.38150	0.02635	0.01223	38	-0.68	036.42	0
036	Sulfur (%)	0229	0.36500	0.01000	0.38150	0.02635	0.01223	38	-0.63	036.04	0
036	Sulfur (%)	0164	0.36500	0.01000	0.38150	0.02635	0.01223	38	-0.63	036.42	0
036	Sulfur (%)	0229	0.36500	0.01000	0.38150	0.02635	0.01223	38	-0.63	036.42	0
036	Sulfur (%)	0171	0.36600	0.00200	0.38150	0.02635	0.01223	38	-0.59	036.42	0
036	Sulfur (%)	0035	0.37000	0.00000	0.38150	0.02635	0.01223	38	-0.44	036.43	0
036	Sulfur (%)	0169	0.37000	0.00000	0.38150	0.02635	0.01223	38	-0.44	036.43	0
036	Sulfur (%)	0083	0.37500	0.01000	0.38150	0.02635	0.01223	38	-0.25	036.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
036	Sulfur (%)	0357	0.37510	0.00420	0.38150	0.02635	0.01223	38	-0.24	036.42	0
036	Sulfur (%)	0407	0.37530	0.00920	0.38150	0.02635	0.01223	38	-0.24	036.42	0
036	Sulfur (%)	0186	0.37600	0.01200	0.38150	0.02635	0.01223	38	-0.21	036.52	0
036	Sulfur (%)	0918	0.37650	0.00500	0.38150	0.02635	0.01223	38	-0.19	036.43	0
036	Sulfur (%)	0187	0.37900	0.00500	0.38150	0.02635	0.01223	38	-0.09	036.42	0
036	Sulfur (%)	0038	0.37950	0.00700	0.38150	0.02635	0.01223	38	-0.08	036.43	0
036	Sulfur (%)	0353	0.38000	0.02000	0.38150	0.02635	0.01223	38	-0.06	036.43	0
036	Sulfur (%)	0668	0.38300	0.01000	0.38150	0.02635	0.01223	38	0.06	036.43	0
036	Sulfur (%)	0202	0.38800	0.00600	0.38150	0.02635	0.01223	38	0.25	036.43	0
036	Sulfur (%)	0226	0.39000	0.02000	0.38150	0.02635	0.01223	38	0.32	036.04	0
036	Sulfur (%)	0560	0.39000	0.01280	0.38150	0.02635	0.01223	38	0.32	036.42	0
036	Sulfur (%)	0148	0.39205	0.01130	0.38150	0.02635	0.01223	38	0.40	036.04	0
036	Sulfur (%)	0186	0.39300	0.00200	0.38150	0.02635	0.01223	38	0.44	036.42	0
036	Sulfur (%)	0035	0.40000	0.02000	0.38150	0.02635	0.01223	38	0.70	036.42	0
036	Sulfur (%)	0278	0.40000	0.02000	0.38150	0.02635	0.01223	38	0.70	036.42	0
036	Sulfur (%)	0242	0.40000	0.02000	0.38150	0.02635	0.01223	38	0.70	036.99	0
036	Sulfur (%)	0098	0.40145	0.00350	0.38150	0.02635	0.01223	38	0.76	036.43	0
036	Sulfur (%)	0218	0.40500	0.01000	0.38150	0.02635	0.01223	38	0.89	036.43	0
036	Sulfur (%)	0708	0.40800	0.00000	0.38150	0.02635	0.01223	38	1.01	036.42	0
036	Sulfur (%)	2089	0.41500	0.05000	0.38150	0.02635	0.01223	38	1.27	036.43	0
036	Sulfur (%)	0553	0.41650	0.01300	0.38150	0.02635	0.01223	38	1.33	036.53	0
036	Sulfur (%)	0032	0.42090	0.02140	0.38150	0.02635	0.01223	38	1.50	036.43	0
036	Sulfur (%)	0510	0.43000	0.00000	0.38150	0.02635	0.01223	38	1.84	036.43	0
036	Sulfur (%)	0294	0.47000	0.02000	0.38150	0.02635	0.01223	38	3.36	036.42	0
036	Sulfur (%)	0910	3,701.5	43.000	0.38150	0.02635	0.01223	38	140437.84	036.52	2
037	Zinc (mg / kg (ppm))	0536	525.94	26.120	698.78	48.161	21.412	94	-3.59	037.31	0
037	Zinc (mg / kg (ppm))	2055	566.00	102.00	698.78	48.161	21.412	94	-2.76	037.53	0
037	Zinc (mg / kg (ppm))	2012	574.74	19.430	698.78	48.161	21.412	94	-2.58	037.41	0
037	Zinc (mg / kg (ppm))	0619	585.50	7.0000	698.78	48.161	21.412	94	-2.35	037.41	0
037	Zinc (mg / kg (ppm))	0164	591.00	14.000	698.78	48.161	21.412	94	-2.24	037.41	0
037	Zinc (mg / kg (ppm))	0047	601.23	0.02000	698.78	48.161	21.412	94	-2.03	037.52	0
037	Zinc (mg / kg (ppm))	0511	603.50	67.000	698.78	48.161	21.412	94	-1.98	037.41	0
037	Zinc (mg / kg (ppm))	2097	609.33	1.2281	698.78	48.161	21.412	94	-1.86	037.31	0
037	Zinc (mg / kg (ppm))	0049	611.73	25.310	698.78	48.161	21.412	94	-1.81	037.41	0
037	Zinc (mg / kg (ppm))	0353	619.00	16.000	698.78	48.161	21.412	94	-1.66	037.43	0
037	Zinc (mg / kg (ppm))	0345	621.50	7.0000	698.78	48.161	21.412	94	-1.60	037.43	0
037	Zinc (mg / kg (ppm))	0148	627.20	32.540	698.78	48.161	21.412	94	-1.49	037.41	0
037	Zinc (mg / kg (ppm))	0590	642.50	5.0000	698.78	48.161	21.412	94	-1.17	037.99	0
037	Zinc (mg / kg (ppm))	0186	645.00	0.00000	698.78	48.161	21.412	94	-1.12	037.42	0
037	Zinc (mg / kg (ppm))	0868	649.04	2.5200	698.78	48.161	21.412	94	-1.03	037.31	0
037	Zinc (mg / kg (ppm))	0038	652.00	6.0000	698.78	48.161	21.412	94	-0.97	037.32	0
037	Zinc (mg / kg (ppm))	0861	654.00	52.000	698.78	48.161	21.412	94	-0.93	037.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
037	Zinc (mg / kg (ppm))	0098	659.85	47.500	698.78	48.161	21.412	94	-0.81	037.41	0
037	Zinc (mg / kg (ppm))	2128	660.57	96.385	698.78	48.161	21.412	94	-0.79	037.32	0
037	Zinc (mg / kg (ppm))	0171	660.80	5.6000	698.78	48.161	21.412	94	-0.79	037.41	0
037	Zinc (mg / kg (ppm))	0278	664.20	8.2000	698.78	48.161	21.412	94	-0.72	037.42	0
037	Zinc (mg / kg (ppm))	0889	665.27	99.489	698.78	48.161	21.412	94	-0.70	037.99	0
037	Zinc (mg / kg (ppm))	0910	667.50	21.000	698.78	48.161	21.412	94	-0.65	037.41	0
037	Zinc (mg / kg (ppm))	0878	668.50	11.000	698.78	48.161	21.412	94	-0.63	037.41	0
037	Zinc (mg / kg (ppm))	0003	669.00	24.000	698.78	48.161	21.412	94	-0.62	037.41	0
037	Zinc (mg / kg (ppm))	0294	670.25	11.700	698.78	48.161	21.412	94	-0.59	037.42	0
037	Zinc (mg / kg (ppm))	2146	674.20	8.2000	698.78	48.161	21.412	94	-0.51	037.51	0
037	Zinc (mg / kg (ppm))	0074	675.00	16.000	698.78	48.161	21.412	94	-0.49	037.41	0
037	Zinc (mg / kg (ppm))	0405	676.00	0.00000	698.78	48.161	21.412	94	-0.47	037.41	0
037	Zinc (mg / kg (ppm))	0037	677.00	4.0000	698.78	48.161	21.412	94	-0.45	037.42	0
037	Zinc (mg / kg (ppm))	2195	678.36	8.1000	698.78	48.161	21.412	94	-0.42	037.31	0
037	Zinc (mg / kg (ppm))	2145	679.72	1.3300	698.78	48.161	21.412	94	-0.40	037.41	0
037	Zinc (mg / kg (ppm))	0358	682.38	6.1300	698.78	48.161	21.412	94	-0.34	037.41	0
037	Zinc (mg / kg (ppm))	2196	683.80	0.00000	698.78	48.161	21.412	94	-0.31	037.31	0
037	Zinc (mg / kg (ppm))	0019	684.35	52.300	698.78	48.161	21.412	94	-0.30	037.41	0
037	Zinc (mg / kg (ppm))	0520	688.00	48.000	698.78	48.161	21.412	94	-0.22	037.41	0
037	Zinc (mg / kg (ppm))	0425	692.00	2.0000	698.78	48.161	21.412	94	-0.14	037.43	0
037	Zinc (mg / kg (ppm))	0045	693.00	82.000	698.78	48.161	21.412	94	-0.12	037.42	0
037	Zinc (mg / kg (ppm))	0610	693.00	4.0000	698.78	48.161	21.412	94	-0.12	037.43	0
037	Zinc (mg / kg (ppm))	0186	693.00	20.000	698.78	48.161	21.412	94	-0.12	037.52	0
037	Zinc (mg / kg (ppm))	0242	693.00	16.000	698.78	48.161	21.412	94	-0.12	037.99	0
037	Zinc (mg / kg (ppm))	0175	695.00	10.000	698.78	48.161	21.412	94	-0.08	037.31	0
037	Zinc (mg / kg (ppm))	2188	695.00	64.000	698.78	48.161	21.412	94	-0.08	037.31	0
037	Zinc (mg / kg (ppm))	0208	696.80	6.2000	698.78	48.161	21.412	94	-0.04	037.41	0
037	Zinc (mg / kg (ppm))	0202	698.55	6.1000	698.78	48.161	21.412	94	0.00	037.43	0
037	Zinc (mg / kg (ppm))	0011	700.38	0.63500	698.78	48.161	21.412	94	0.03	037.41	0
037	Zinc (mg / kg (ppm))	0693	702.46	42.886	698.78	48.161	21.412	94	0.08	037.42	0
037	Zinc (mg / kg (ppm))	0407	702.98	0.92600	698.78	48.161	21.412	94	0.09	037.41	0
037	Zinc (mg / kg (ppm))	0939	703.43	17.670	698.78	48.161	21.412	94	0.10	037.31	0
037	Zinc (mg / kg (ppm))	0265	703.50	15.000	698.78	48.161	21.412	94	0.10	037.42	0
037	Zinc (mg / kg (ppm))	0033	704.00	0.00000	698.78	48.161	21.412	94	0.11	037.43	0
037	Zinc (mg / kg (ppm))	0668	708.00	22.000	698.78	48.161	21.412	94	0.19	037.43	0
037	Zinc (mg / kg (ppm))	0968	708.00	2.0000	698.78	48.161	21.412	94	0.19	037.43	0
037	Zinc (mg / kg (ppm))	0226	708.50	15.000	698.78	48.161	21.412	94	0.20	037.41	0
037	Zinc (mg / kg (ppm))	2022	709.66	0.29000	698.78	48.161	21.412	94	0.23	037.31	0
037	Zinc (mg / kg (ppm))	0187	710.01	5.1900	698.78	48.161	21.412	94	0.23	037.42	0
037	Zinc (mg / kg (ppm))	0354	710.35	6.5100	698.78	48.161	21.412	94	0.24	037.31	0
037	Zinc (mg / kg (ppm))	0598	712.55	35.500	698.78	48.161	21.412	94	0.29	037.41	0
037	Zinc (mg / kg (ppm))	0918	713.15	17.630	698.78	48.161	21.412	94	0.30	037.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
037	Zinc (mg / kg (ppm))	0035	722.00	44.000	698.78	48.161	21.412	94	0.48	037.43	0
037	Zinc (mg / kg (ppm))	0098	722.90	20.200	698.78	48.161	21.412	94	0.50	037.43	0
037	Zinc (mg / kg (ppm))	0560	724.40	7.4000	698.78	48.161	21.412	94	0.53	037.42	0
037	Zinc (mg / kg (ppm))	0357	724.42	12.169	698.78	48.161	21.412	94	0.53	037.42	0
037	Zinc (mg / kg (ppm))	0013	725.00	6.0000	698.78	48.161	21.412	94	0.54	037.43	0
037	Zinc (mg / kg (ppm))	0218	725.00	30.000	698.78	48.161	21.412	94	0.54	037.43	0
037	Zinc (mg / kg (ppm))	0689	725.15	19.500	698.78	48.161	21.412	94	0.55	037.31	0
037	Zinc (mg / kg (ppm))	0042	726.00	60.000	698.78	48.161	21.412	94	0.57	037.42	0
037	Zinc (mg / kg (ppm))	2144	726.09	20.150	698.78	48.161	21.412	94	0.57	037.99	0
037	Zinc (mg / kg (ppm))	0413	728.00	0.00000	698.78	48.161	21.412	94	0.61	037.42	0
037	Zinc (mg / kg (ppm))	0229	728.32	50.510	698.78	48.161	21.412	94	0.61	037.41	0
037	Zinc (mg / kg (ppm))	0017	730.16	27.168	698.78	48.161	21.412	94	0.65	037.43	0
037	Zinc (mg / kg (ppm))	0870	731.30	50.590	698.78	48.161	21.412	94	0.68	037.42	0
037	Zinc (mg / kg (ppm))	0510	735.00	2.0000	698.78	48.161	21.412	94	0.75	037.43	0
037	Zinc (mg / kg (ppm))	0675	735.73	19.180	698.78	48.161	21.412	94	0.77	037.43	0
037	Zinc (mg / kg (ppm))	0948	735.83	1.7520	698.78	48.161	21.412	94	0.77	037.33	0
037	Zinc (mg / kg (ppm))	0034	737.15	26.700	698.78	48.161	21.412	94	0.80	037.43	0
037	Zinc (mg / kg (ppm))	0208	740.00	40.000	698.78	48.161	21.412	94	0.86	037.31	0
037	Zinc (mg / kg (ppm))	0964	740.35	56.321	698.78	48.161	21.412	94	0.86	037.43	0
037	Zinc (mg / kg (ppm))	0505	740.60	7.8000	698.78	48.161	21.412	94	0.87	037.43	0
037	Zinc (mg / kg (ppm))	0682	741.48	0.00000	698.78	48.161	21.412	94	0.89	037.43	0
037	Zinc (mg / kg (ppm))	0572	743.00	32.000	698.78	48.161	21.412	94	0.92	037.53	0
037	Zinc (mg / kg (ppm))	0001	743.25	15.700	698.78	48.161	21.412	94	0.92	037.31	0
037	Zinc (mg / kg (ppm))	0874	753.50	19.000	698.78	48.161	21.412	94	1.14	037.31	0
037	Zinc (mg / kg (ppm))	0010	755.00	10.000	698.78	48.161	21.412	94	1.17	037.33	0
037	Zinc (mg / kg (ppm))	0032	755.50	5.0000	698.78	48.161	21.412	94	1.18	037.43	0
037	Zinc (mg / kg (ppm))	0009	756.50	19.000	698.78	48.161	21.412	94	1.20	037.42	0
037	Zinc (mg / kg (ppm))	0190	757.53	19.360	698.78	48.161	21.412	94	1.22	037.42	0
037	Zinc (mg / kg (ppm))	2053	763.00	42.000	698.78	48.161	21.412	94	1.33	037.42	0
037	Zinc (mg / kg (ppm))	0035	765.50	3.0000	698.78	48.161	21.412	94	1.39	037.42	0
037	Zinc (mg / kg (ppm))	0083	780.00	12.000	698.78	48.161	21.412	94	1.69	037.43	0
037	Zinc (mg / kg (ppm))	0169	780.50	9.0000	698.78	48.161	21.412	94	1.70	037.43	0
037	Zinc (mg / kg (ppm))	0100	783.00	8.0000	698.78	48.161	21.412	94	1.75	037.99	0
037	Zinc (mg / kg (ppm))	0027	786.81	57.559	698.78	48.161	21.412	94	1.83	037.43	0
037	Zinc (mg / kg (ppm))	0553	796.00	16.000	698.78	48.161	21.412	94	2.02	037.53	0
037	Zinc (mg / kg (ppm))	2089	739.32	237.71	698.78	48.161	21.412	94	0.84	037.43	1
038	Molybdenum (mg / kg (ppm))	0910	2.6500	0.12000	4.0359	0.96675	0.20295	22	-1.43	038.52	0
038	Molybdenum (mg / kg (ppm))	2055	2.9150	1.0500	4.0359	0.96675	0.20295	22	-1.16	038.53	0
038	Molybdenum (mg / kg (ppm))	0345	2.9500	0.20000	4.0359	0.96675	0.20295	22	-1.12	038.43	0
038	Molybdenum (mg / kg (ppm))	0186	3.0000	0.00000	4.0359	0.96675	0.20295	22	-1.07	038.52	0
038	Molybdenum (mg / kg (ppm))	0171	3.1485	0.03700	4.0359	0.96675	0.20295	22	-0.92	038.41	0
038	Molybdenum (mg / kg (ppm))	0918	3.3312	0.11150	4.0359	0.96675	0.20295	22	-0.73	038.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
038	Molybdenum (mg / kg (ppm))	0353	3.3391	0.23940	4.0359	0.96675	0.20295	22	-0.72	038.43	0
038	Molybdenum (mg / kg (ppm))	0226	3.6300	0.26000	4.0359	0.96675	0.20295	22	-0.42	038.41	0
038	Molybdenum (mg / kg (ppm))	0278	3.7450	0.33000	4.0359	0.96675	0.20295	22	-0.30	038.42	0
038	Molybdenum (mg / kg (ppm))	0098	3.9020	0.13400	4.0359	0.96675	0.20295	22	-0.14	038.53	0
038	Molybdenum (mg / kg (ppm))	0169	3.9550	0.01000	4.0359	0.96675	0.20295	22	-0.08	038.43	0
038	Molybdenum (mg / kg (ppm))	0218	4.0000	0.00000	4.0359	0.96675	0.20295	22	-0.04	038.43	0
038	Molybdenum (mg / kg (ppm))	0693	4.2695	0.19300	4.0359	0.96675	0.20295	22	0.24	038.42	0
038	Molybdenum (mg / kg (ppm))	0265	4.4050	0.29000	4.0359	0.96675	0.20295	22	0.38	038.42	0
038	Molybdenum (mg / kg (ppm))	0083	4.4500	0.30000	4.0359	0.96675	0.20295	22	0.43	038.43	0
038	Molybdenum (mg / kg (ppm))	0553	4.5600	0.00000	4.0359	0.96675	0.20295	22	0.54	038.53	0
038	Molybdenum (mg / kg (ppm))	2053	4.6550	0.43000	4.0359	0.96675	0.20295	22	0.64	038.42	0
038	Molybdenum (mg / kg (ppm))	0510	4.8000	0.00000	4.0359	0.96675	0.20295	22	0.79	038.43	0
038	Molybdenum (mg / kg (ppm))	0045	5.0900	0.44000	4.0359	0.96675	0.20295	22	1.09	038.42	0
038	Molybdenum (mg / kg (ppm))	0218	5.2500	0.30000	4.0359	0.96675	0.20295	22	1.26	038.53	0
038	Molybdenum (mg / kg (ppm))	0032	5.2590	0.01800	4.0359	0.96675	0.20295	22	1.27	038.43	0
038	Molybdenum (mg / kg (ppm))	0148	9.1990	0.00200	4.0359	0.96675	0.20295	22	5.34	038.41	0
038	Molybdenum (mg / kg (ppm))	0038	5.0125	1.1650	4.0359	0.96675	0.20295	22	1.01	038.43	1
040	Barium (mg / kg (ppm))	0560	8.4400	0.02000				2		040.42	0
040	Barium (mg / kg (ppm))	0918	8.7502	0.24740				2		040.53	0
041	Vanadium (mg / kg (ppm))	0553	3.8750	0.15000				3		041.53	0
041	Vanadium (mg / kg (ppm))	0218	3.9000	0.20000				3		041.43	0
041	Vanadium (mg / kg (ppm))	0218	4.6500	0.10000				3		041.53	0
042	Chloride (%)	0948	2.1350	0.01000				2		042.00	0
042	Chloride (%)	0536	2.3850	0.01000				2		042.99	0
095	Methoprene (mg/kg (ppm))	0675	4.6800	1.2800				3		095.01	0
095	Methoprene (mg/kg (ppm))	2053	11.500	1.0000				3		095.99	0
095	Methoprene (mg/kg (ppm))	0019	17.695	0.53000				3		095.01	0
101	Choline Chloride (mg / kg (ppm))	0227	1,155.0	50.000				1		101.02	0
102	Niacin (mg / kg (ppm))	0227	55.850	4.9000				1		102.01	0
103	Pantothenic Acid (mg / kg (ppm))	0227	5.3850	0.47000				1		103.01	0
104	Riboflavin (mg / kg (ppm))	0910	1.3850	0.57000				2		104.03	0
104	Riboflavin (mg / kg (ppm))	0227	6.6350	0.87000				2		104.00	0
105	Thiamine (mg / kg (ppm))	0910	1.1500	0.32000				2		105.00	0
105	Thiamine (mg / kg (ppm))	0227	1.3400	0.12000				2		105.01	0
106	Vitamin A (KU / kg)	0169	7.1500	0.30000	24.826	6.4085	1.9500	25	-2.76	106.02	0
106	Vitamin A (KU / kg)	0861	17.256	2.3580	24.826	6.4085	1.9500	25	-1.18	106.02	0
106	Vitamin A (KU / kg)	2053	18.950	1.1000	24.826	6.4085	1.9500	25	-0.92	106.02	0
106	Vitamin A (KU / kg)	0941	19.711	0.94700	24.826	6.4085	1.9500	25	-0.80	106.02	0
106	Vitamin A (KU / kg)	0689	19.850	0.30000	24.826	6.4085	1.9500	25	-0.78	106.02	0
106	Vitamin A (KU / kg)	0676	20.070	0.98000	24.826	6.4085	1.9500	25	-0.74	106.02	0
106	Vitamin A (KU / kg)	0027	20.709	1.7730	24.826	6.4085	1.9500	25	-0.64	106.02	0
106	Vitamin A (KU / kg)	0904	21.005	1.2700	24.826	6.4085	1.9500	25	-0.60	106.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
106	Vitamin A (KU / kg)	0017	21.270	4.1600	24.826	6.4085	1.9500	25	-0.55	106.02	0
106	Vitamin A (KU / kg)	0035	21.725	1.0300	24.826	6.4085	1.9500	25	-0.48	106.02	0
106	Vitamin A (KU / kg)	0148	22.460	0.00000	24.826	6.4085	1.9500	25	-0.37	106.02	0
106	Vitamin A (KU / kg)	0905	23.230	0.92900	24.826	6.4085	1.9500	25	-0.25	106.02	0
106	Vitamin A (KU / kg)	0910	23.950	0.10000	24.826	6.4085	1.9500	25	-0.14	106.02	0
106	Vitamin A (KU / kg)	2103	25.200	1.1600	24.826	6.4085	1.9500	25	0.06	106.02	0
106	Vitamin A (KU / kg)	0160	25.200	1.8000	24.826	6.4085	1.9500	25	0.06	106.99	0
106	Vitamin A (KU / kg)	0004	25.780	3.7400	24.826	6.4085	1.9500	25	0.15	106.02	0
106	Vitamin A (KU / kg)	0227	26.500	7.0000	24.826	6.4085	1.9500	25	0.26	106.02	0
106	Vitamin A (KU / kg)	0014	26.900	0.80000	24.826	6.4085	1.9500	25	0.32	106.02	0
106	Vitamin A (KU / kg)	2022	27.000	2.0000	24.826	6.4085	1.9500	25	0.34	106.02	0
106	Vitamin A (KU / kg)	2144	30.315	0.01000	24.826	6.4085	1.9500	25	0.86	106.02	0
106	Vitamin A (KU / kg)	0098	30.650	7.5000	24.826	6.4085	1.9500	25	0.91	106.01	0
106	Vitamin A (KU / kg)	0010	34.400	6.6000	24.826	6.4085	1.9500	25	1.49	106.02	0
106	Vitamin A (KU / kg)	0675	59.440	2.2600	24.826	6.4085	1.9500	25	5.40	106.02	0
106	Vitamin A (KU / kg)	0019	60.987	0.63300	24.826	6.4085	1.9500	25	5.64	106.00	0
106	Vitamin A (KU / kg)	0638	84.550	0.00000	24.826	6.4085	1.9500	25	9.32	106.02	0
107	Vitamin B12 (µg / kg (ppb))	0227	29.500	9.0000				1		107.00	0
108	Vitamin D3 (KU / kg)	0169	1.8500	0.30000	8.3442	5.9834	1.0387	7	-1.09	108.01	0
108	Vitamin D3 (KU / kg)	0160	2.3400	0.80000	8.3442	5.9834	1.0387	7	-1.00	108.01	0
108	Vitamin D3 (KU / kg)	0676	4.9545	0.10100	8.3442	5.9834	1.0387	7	-0.57	108.02	0
108	Vitamin D3 (KU / kg)	0904	9.8400	0.68000	8.3442	5.9834	1.0387	7	0.25	108.99	0
108	Vitamin D3 (KU / kg)	0910	11.305	1.1500	8.3442	5.9834	1.0387	7	0.49	108.02	0
108	Vitamin D3 (KU / kg)	0861	13.360	4.2400	8.3442	5.9834	1.0387	7	0.84	108.02	0
108	Vitamin D3 (KU / kg)	0675	14.760	0.00000	8.3442	5.9834	1.0387	7	1.07	108.02	0
108	Vitamin D3 (KU / kg)	0227	11,725	5,550.0	8.3442	5.9834	1.0387	7	1958.20	108.02	2
109	Vitamin E (IU / kg)	0638	5.8000	0.00000	41.027	11.318	2.0276	17	-3.11	109.02	0
109	Vitamin E (IU / kg)	0169	8.2500	0.10000	41.027	11.318	2.0276	17	-2.90	109.02	0
109	Vitamin E (IU / kg)	0171	19.525	0.05000	41.027	11.318	2.0276	17	-1.90	109.99	0
109	Vitamin E (IU / kg)	2022	37.500	1.0000	41.027	11.318	2.0276	17	-0.31	109.02	0
109	Vitamin E (IU / kg)	0941	37.720	1.8600	41.027	11.318	2.0276	17	-0.29	109.02	0
109	Vitamin E (IU / kg)	2103	38.900	1.6000	41.027	11.318	2.0276	17	-0.19	109.02	0
109	Vitamin E (IU / kg)	0910	39.650	0.90000	41.027	11.318	2.0276	17	-0.12	109.02	0
109	Vitamin E (IU / kg)	0208	40.816	1.6400	41.027	11.318	2.0276	17	-0.02	109.02	0
109	Vitamin E (IU / kg)	0861	41.000	8.0000	41.027	11.318	2.0276	17	0.00	109.02	0
109	Vitamin E (IU / kg)	0148	41.280	0.00000	41.027	11.318	2.0276	17	0.02	109.02	0
109	Vitamin E (IU / kg)	0905	41.500	5.0000	41.027	11.318	2.0276	17	0.04	109.02	0
109	Vitamin E (IU / kg)	0098	47.300	6.6000	41.027	11.318	2.0276	17	0.55	109.02	0
109	Vitamin E (IU / kg)	0676	47.800	0.20000	41.027	11.318	2.0276	17	0.60	109.02	0
109	Vitamin E (IU / kg)	0227	49.950	0.30000	41.027	11.318	2.0276	17	0.79	109.02	0
109	Vitamin E (IU / kg)	0904	50.000	2.5200	41.027	11.318	2.0276	17	0.79	109.99	0
109	Vitamin E (IU / kg)	0675	53.890	0.40000	41.027	11.318	2.0276	17	1.14	109.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
109	Vitamin E (IU / kg)	0160	61.750	4.3000	41.027	11.318	2.0276	17	1.83	109.99	0
112	Pyridoxine (µg / g)	0227	3.8800	0.26000				1		112.01	0
113	Folic Acid (mg / kg (ppm))	0227	0.50700	0.13400				1		113.01	0
114	Biotin (mg / kg (ppm))	0227	0.19900	0.01800				1		114.01	0
115	Non Protein N (NPN) (%)	0098	0.74000	0.04000				1		115.00	0
118	Peroxide value (meq/kg)	2081	20.000	6.6000				1		118.99	0
120	Alanine (%)	0353	0.55000	0.00000	0.64819	0.03289	0.01304	31	-2.99	120.00	0
120	Alanine (%)	0676	0.59000	0.02000	0.64819	0.03289	0.01304	31	-1.77	120.05	0
120	Alanine (%)	0407	0.60400	0.00800	0.64819	0.03289	0.01304	31	-1.34	120.05	0
120	Alanine (%)	0870	0.61100	0.03520	0.64819	0.03289	0.01304	31	-1.13	120.00	0
120	Alanine (%)	0675	0.62000	0.02000	0.64819	0.03289	0.01304	31	-0.86	120.00	0
120	Alanine (%)	0910	0.62000	0.04000	0.64819	0.03289	0.01304	31	-0.86	120.00	0
120	Alanine (%)	0148	0.62000	0.00000	0.64819	0.03289	0.01304	31	-0.86	120.05	0
120	Alanine (%)	0626	0.63050	0.01300	0.64819	0.03289	0.01304	31	-0.54	120.05	0
120	Alanine (%)	0878	0.63100	0.00000	0.64819	0.03289	0.01304	31	-0.52	120.00	0
120	Alanine (%)	0354	0.63700	0.00200	0.64819	0.03289	0.01304	31	-0.34	120.00	0
120	Alanine (%)	0504	0.64000	0.02000	0.64819	0.03289	0.01304	31	-0.25	120.00	0
120	Alanine (%)	0571	0.64000	0.00600	0.64819	0.03289	0.01304	31	-0.25	120.00	0
120	Alanine (%)	0644	0.64050	0.00900	0.64819	0.03289	0.01304	31	-0.23	120.00	0
120	Alanine (%)	0868	0.64150	0.02100	0.64819	0.03289	0.01304	31	-0.20	120.00	0
120	Alanine (%)	0652	0.64500	0.01000	0.64819	0.03289	0.01304	31	-0.10	120.00	0
120	Alanine (%)	0684	0.64550	0.05900	0.64819	0.03289	0.01304	31	-0.08	120.00	0
120	Alanine (%)	0208	0.65000	0.02000	0.64819	0.03289	0.01304	31	0.06	120.00	0
120	Alanine (%)	0227	0.65000	0.00000	0.64819	0.03289	0.01304	31	0.06	120.00	0
120	Alanine (%)	0098	0.65150	0.00700	0.64819	0.03289	0.01304	31	0.10	120.02	0
120	Alanine (%)	0662	0.65435	0.01970	0.64819	0.03289	0.01304	31	0.19	120.00	0
120	Alanine (%)	0872	0.65450	0.00500	0.64819	0.03289	0.01304	31	0.19	120.00	0
120	Alanine (%)	0918	0.65475	0.00950	0.64819	0.03289	0.01304	31	0.20	120.00	0
120	Alanine (%)	2059	0.65550	0.00900	0.64819	0.03289	0.01304	31	0.22	120.00	0
120	Alanine (%)	0968	0.66350	0.00900	0.64819	0.03289	0.01304	31	0.47	120.00	0
120	Alanine (%)	0859	0.67350	0.00100	0.64819	0.03289	0.01304	31	0.77	120.00	0
120	Alanine (%)	2196	0.68500	0.00000	0.64819	0.03289	0.01304	31	1.12	120.05	0
120	Alanine (%)	0619	0.69000	0.01600	0.64819	0.03289	0.01304	31	1.27	120.00	0
120	Alanine (%)	0941	0.69500	0.00200	0.64819	0.03289	0.01304	31	1.42	120.00	0
120	Alanine (%)	0226	0.72550	0.03280	0.64819	0.03289	0.01304	31	2.35	120.00	0
120	Alanine (%)	0682	0.81200	0.00000	0.64819	0.03289	0.01304	31	4.98	120.00	0
120	Alanine (%)	2188	0.86500	0.01000	0.64819	0.03289	0.01304	31	6.59	120.05	0
120	Alanine (%)	0853	0.79500	0.09000	0.64819	0.03289	0.01304	31	4.46	120.05	1
121	Arginine (%)	0353	0.39000	0.02000	0.47373	0.02362	0.01363	32	-3.55	121.00	0
121	Arginine (%)	0407	0.43100	0.02400	0.47373	0.02362	0.01363	32	-1.81	121.05	0
121	Arginine (%)	0354	0.44400	0.00400	0.47373	0.02362	0.01363	32	-1.26	121.00	0
121	Arginine (%)	0227	0.45000	0.04000	0.47373	0.02362	0.01363	32	-1.00	121.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
121	Arginine (%)	0675	0.45000	0.02000	0.47373	0.02362	0.01363	32	-1.00	121.00	0
121	Arginine (%)	0098	0.45050	0.01300	0.47373	0.02362	0.01363	32	-0.98	121.02	0
121	Arginine (%)	0504	0.45500	0.01000	0.47373	0.02362	0.01363	32	-0.79	121.00	0
121	Arginine (%)	0684	0.45500	0.00800	0.47373	0.02362	0.01363	32	-0.79	121.00	0
121	Arginine (%)	0853	0.46000	0.04000	0.47373	0.02362	0.01363	32	-0.58	121.05	0
121	Arginine (%)	0226	0.46310	0.03960	0.47373	0.02362	0.01363	32	-0.45	121.00	0
121	Arginine (%)	0910	0.46500	0.03000	0.47373	0.02362	0.01363	32	-0.37	121.00	0
121	Arginine (%)	0148	0.46500	0.01000	0.47373	0.02362	0.01363	32	-0.37	121.05	0
121	Arginine (%)	0676	0.46500	0.03000	0.47373	0.02362	0.01363	32	-0.37	121.05	0
121	Arginine (%)	0878	0.47000	0.00200	0.47373	0.02362	0.01363	32	-0.16	121.00	0
121	Arginine (%)	0870	0.47195	0.00890	0.47373	0.02362	0.01363	32	-0.08	121.00	0
121	Arginine (%)	0918	0.47225	0.01450	0.47373	0.02362	0.01363	32	-0.06	121.00	0
121	Arginine (%)	0208	0.47500	0.01000	0.47373	0.02362	0.01363	32	0.05	121.00	0
121	Arginine (%)	0652	0.47500	0.01000	0.47373	0.02362	0.01363	32	0.05	121.00	0
121	Arginine (%)	0868	0.47500	0.00200	0.47373	0.02362	0.01363	32	0.05	121.00	0
121	Arginine (%)	0872	0.47550	0.00700	0.47373	0.02362	0.01363	32	0.07	121.00	0
121	Arginine (%)	0662	0.47805	0.01210	0.47373	0.02362	0.01363	32	0.18	121.00	0
121	Arginine (%)	2059	0.48400	0.00600	0.47373	0.02362	0.01363	32	0.43	121.00	0
121	Arginine (%)	0626	0.48450	0.01100	0.47373	0.02362	0.01363	32	0.46	121.05	0
121	Arginine (%)	0571	0.48700	0.00600	0.47373	0.02362	0.01363	32	0.56	121.00	0
121	Arginine (%)	0968	0.48850	0.00100	0.47373	0.02362	0.01363	32	0.63	121.00	0
121	Arginine (%)	0644	0.49100	0.00600	0.47373	0.02362	0.01363	32	0.73	121.00	0
121	Arginine (%)	0941	0.49600	0.00400	0.47373	0.02362	0.01363	32	0.94	121.00	0
121	Arginine (%)	2188	0.50000	0.02000	0.47373	0.02362	0.01363	32	1.11	121.05	0
121	Arginine (%)	2196	0.50900	0.00000	0.47373	0.02362	0.01363	32	1.49	121.05	0
121	Arginine (%)	0859	0.51650	0.00700	0.47373	0.02362	0.01363	32	1.81	121.00	0
121	Arginine (%)	0619	0.51900	0.02000	0.47373	0.02362	0.01363	32	1.92	121.00	0
121	Arginine (%)	0682	0.56500	0.00000	0.47373	0.02362	0.01363	32	3.86	121.00	0
122	Aspartic (%)	0353	0.68500	0.01000	0.78274	0.03356	0.01267	31	-2.91	122.00	0
122	Aspartic (%)	0676	0.70500	0.01000	0.78274	0.03356	0.01267	31	-2.32	122.05	0
122	Aspartic (%)	0675	0.71500	0.03000	0.78274	0.03356	0.01267	31	-2.02	122.00	0
122	Aspartic (%)	0684	0.74000	0.01600	0.78274	0.03356	0.01267	31	-1.27	122.00	0
122	Aspartic (%)	0354	0.75050	0.00100	0.78274	0.03356	0.01267	31	-0.96	122.00	0
122	Aspartic (%)	0626	0.75350	0.00100	0.78274	0.03356	0.01267	31	-0.87	122.05	0
122	Aspartic (%)	0504	0.76000	0.00000	0.78274	0.03356	0.01267	31	-0.68	122.00	0
122	Aspartic (%)	0227	0.76500	0.01000	0.78274	0.03356	0.01267	31	-0.53	122.00	0
122	Aspartic (%)	0407	0.76700	0.06800	0.78274	0.03356	0.01267	31	-0.47	122.05	0
122	Aspartic (%)	0878	0.77450	0.00700	0.78274	0.03356	0.01267	31	-0.25	122.00	0
122	Aspartic (%)	0868	0.77500	0.00600	0.78274	0.03356	0.01267	31	-0.23	122.00	0
122	Aspartic (%)	0910	0.77500	0.01000	0.78274	0.03356	0.01267	31	-0.23	122.00	0
122	Aspartic (%)	2188	0.77500	0.03000	0.78274	0.03356	0.01267	31	-0.23	122.05	0
122	Aspartic (%)	0644	0.77550	0.02300	0.78274	0.03356	0.01267	31	-0.22	122.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
122	Aspartic (%)	0662	0.77565	0.01870	0.78274	0.03356	0.01267	31	-0.21	122.00	0
122	Aspartic (%)	0872	0.77950	0.00300	0.78274	0.03356	0.01267	31	-0.10	122.00	0
122	Aspartic (%)	0571	0.78100	0.00200	0.78274	0.03356	0.01267	31	-0.05	122.00	0
122	Aspartic (%)	2059	0.78600	0.00800	0.78274	0.03356	0.01267	31	0.10	122.00	0
122	Aspartic (%)	0870	0.78760	0.00620	0.78274	0.03356	0.01267	31	0.14	122.00	0
122	Aspartic (%)	0098	0.78800	0.00000	0.78274	0.03356	0.01267	31	0.16	122.02	0
122	Aspartic (%)	0208	0.79000	0.00000	0.78274	0.03356	0.01267	31	0.22	122.00	0
122	Aspartic (%)	0652	0.79000	0.00000	0.78274	0.03356	0.01267	31	0.22	122.00	0
122	Aspartic (%)	2196	0.80500	0.00000	0.78274	0.03356	0.01267	31	0.66	122.05	0
122	Aspartic (%)	0148	0.81000	0.02000	0.78274	0.03356	0.01267	31	0.81	122.05	0
122	Aspartic (%)	0941	0.81100	0.02200	0.78274	0.03356	0.01267	31	0.84	122.00	0
122	Aspartic (%)	0968	0.81350	0.00300	0.78274	0.03356	0.01267	31	0.92	122.00	0
122	Aspartic (%)	0918	0.81865	0.00950	0.78274	0.03356	0.01267	31	1.07	122.00	0
122	Aspartic (%)	0859	0.82150	0.01300	0.78274	0.03356	0.01267	31	1.16	122.00	0
122	Aspartic (%)	0226	0.85890	0.04740	0.78274	0.03356	0.01267	31	2.27	122.00	0
122	Aspartic (%)	0619	0.88600	0.01800	0.78274	0.03356	0.01267	31	3.08	122.00	0
122	Aspartic (%)	0682	0.92800	0.00000	0.78274	0.03356	0.01267	31	4.33	122.00	0
122	Aspartic (%)	0853	0.98000	0.10000	0.78274	0.03356	0.01267	31	5.88	122.05	1
124	Cysteine/Cystine (%)	0853	0.01000	0.00000	0.20831	0.02659	0.00762	29	-7.46	124.05	0
124	Cysteine/Cystine (%)	0684	0.02900	0.00200	0.20831	0.02659	0.00762	29	-6.74	124.00	0
124	Cysteine/Cystine (%)	0941	0.16350	0.00300	0.20831	0.02659	0.00762	29	-1.69	124.00	0
124	Cysteine/Cystine (%)	0353	0.16500	0.01000	0.20831	0.02659	0.00762	29	-1.63	124.00	0
124	Cysteine/Cystine (%)	0870	0.17880	0.00780	0.20831	0.02659	0.00762	29	-1.11	124.00	0
124	Cysteine/Cystine (%)	0662	0.18700	0.00260	0.20831	0.02659	0.00762	29	-0.80	124.00	0
124	Cysteine/Cystine (%)	0910	0.19000	0.02000	0.20831	0.02659	0.00762	29	-0.69	124.00	0
124	Cysteine/Cystine (%)	0676	0.19500	0.01000	0.20831	0.02659	0.00762	29	-0.50	124.05	0
124	Cysteine/Cystine (%)	0227	0.20000	0.02000	0.20831	0.02659	0.00762	29	-0.31	124.00	0
124	Cysteine/Cystine (%)	0504	0.20500	0.01000	0.20831	0.02659	0.00762	29	-0.12	124.00	0
124	Cysteine/Cystine (%)	0652	0.20500	0.01000	0.20831	0.02659	0.00762	29	-0.12	124.00	0
124	Cysteine/Cystine (%)	0878	0.20650	0.00700	0.20831	0.02659	0.00762	29	-0.07	124.00	0
124	Cysteine/Cystine (%)	2059	0.20750	0.00300	0.20831	0.02659	0.00762	29	-0.03	124.00	0
124	Cysteine/Cystine (%)	0968	0.21000	0.00200	0.20831	0.02659	0.00762	29	0.06	124.00	0
124	Cysteine/Cystine (%)	0918	0.21000	0.02000	0.20831	0.02659	0.00762	29	0.06	124.00	0
124	Cysteine/Cystine (%)	0644	0.21300	0.01200	0.20831	0.02659	0.00762	29	0.18	124.00	0
124	Cysteine/Cystine (%)	0571	0.21350	0.00900	0.20831	0.02659	0.00762	29	0.20	124.00	0
124	Cysteine/Cystine (%)	2196	0.21400	0.00000	0.20831	0.02659	0.00762	29	0.21	124.05	0
124	Cysteine/Cystine (%)	0872	0.21500	0.01400	0.20831	0.02659	0.00762	29	0.25	124.00	0
124	Cysteine/Cystine (%)	0098	0.21600	0.00000	0.20831	0.02659	0.00762	29	0.29	124.02	0
124	Cysteine/Cystine (%)	0354	0.22000	0.00200	0.20831	0.02659	0.00762	29	0.44	124.00	0
124	Cysteine/Cystine (%)	0682	0.22000	0.00000	0.20831	0.02659	0.00762	29	0.44	124.00	0
124	Cysteine/Cystine (%)	0859	0.22250	0.00300	0.20831	0.02659	0.00762	29	0.53	124.00	0
124	Cysteine/Cystine (%)	0868	0.22450	0.00100	0.20831	0.02659	0.00762	29	0.61	124.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
124	Cysteine/Cystine (%)	0208	0.23050	0.02500	0.20831	0.02659	0.00762	29	0.83	124.00	0
124	Cysteine/Cystine (%)	0226	0.24055	0.00850	0.20831	0.02659	0.00762	29	1.21	124.00	0
124	Cysteine/Cystine (%)	0619	0.24650	0.00900	0.20831	0.02659	0.00762	29	1.44	124.00	0
124	Cysteine/Cystine (%)	0675	0.25500	0.01000	0.20831	0.02659	0.00762	29	1.76	124.00	0
124	Cysteine/Cystine (%)	2188	0.30000	0.00000	0.20831	0.02659	0.00762	29	3.45	124.05	0
124	Cysteine/Cystine (%)	0148	0.23500	0.05000	0.20831	0.02659	0.00762	29	1.00	124.05	1
124	Cysteine/Cystine (%)	0407	0.23650	0.05100	0.20831	0.02659	0.00762	29	1.06	124.05	1
125	Glutamic (%)	0676	1.4400	0.02000	1.6212	0.09212	0.02309	31	-1.97	125.05	0
125	Glutamic (%)	0353	1.4650	0.01000	1.6212	0.09212	0.02309	31	-1.70	125.00	0
125	Glutamic (%)	0675	1.5200	0.04000	1.6212	0.09212	0.02309	31	-1.10	125.00	0
125	Glutamic (%)	2188	1.5200	0.00000	1.6212	0.09212	0.02309	31	-1.10	125.05	0
125	Glutamic (%)	0407	1.5345	0.11500	1.6212	0.09212	0.02309	31	-0.94	125.05	0
125	Glutamic (%)	0504	1.5400	0.04000	1.6212	0.09212	0.02309	31	-0.88	125.00	0
125	Glutamic (%)	0684	1.5515	0.06900	1.6212	0.09212	0.02309	31	-0.76	125.00	0
125	Glutamic (%)	0227	1.5600	0.02000	1.6212	0.09212	0.02309	31	-0.66	125.00	0
125	Glutamic (%)	0644	1.5770	0.00200	1.6212	0.09212	0.02309	31	-0.48	125.00	0
125	Glutamic (%)	0571	1.5820	0.02000	1.6212	0.09212	0.02309	31	-0.43	125.00	0
125	Glutamic (%)	2059	1.5820	0.01600	1.6212	0.09212	0.02309	31	-0.43	125.00	0
125	Glutamic (%)	0354	1.5920	0.00200	1.6212	0.09212	0.02309	31	-0.32	125.00	0
125	Glutamic (%)	0872	1.5995	0.00900	1.6212	0.09212	0.02309	31	-0.24	125.00	0
125	Glutamic (%)	0098	1.6020	0.01400	1.6212	0.09212	0.02309	31	-0.21	125.02	0
125	Glutamic (%)	0878	1.6020	0.00400	1.6212	0.09212	0.02309	31	-0.21	125.00	0
125	Glutamic (%)	0910	1.6050	0.03000	1.6212	0.09212	0.02309	31	-0.18	125.00	0
125	Glutamic (%)	0868	1.6120	0.01000	1.6212	0.09212	0.02309	31	-0.10	125.00	0
125	Glutamic (%)	0626	1.6215	0.04500	1.6212	0.09212	0.02309	31	0.00	125.05	0
125	Glutamic (%)	0652	1.6250	0.01000	1.6212	0.09212	0.02309	31	0.04	125.00	0
125	Glutamic (%)	0968	1.6310	0.00200	1.6212	0.09212	0.02309	31	0.11	125.00	0
125	Glutamic (%)	0662	1.6433	0.03680	1.6212	0.09212	0.02309	31	0.24	125.00	0
125	Glutamic (%)	2196	1.6630	0.00000	1.6212	0.09212	0.02309	31	0.45	125.05	0
125	Glutamic (%)	0918	1.6679	0.04660	1.6212	0.09212	0.02309	31	0.51	125.00	0
125	Glutamic (%)	0941	1.6825	0.00900	1.6212	0.09212	0.02309	31	0.67	125.00	0
125	Glutamic (%)	0870	1.7073	0.00200	1.6212	0.09212	0.02309	31	0.94	125.00	0
125	Glutamic (%)	0148	1.7150	0.01000	1.6212	0.09212	0.02309	31	1.02	125.05	0
125	Glutamic (%)	0859	1.7310	0.02800	1.6212	0.09212	0.02309	31	1.19	125.00	0
125	Glutamic (%)	0208	1.7450	0.01000	1.6212	0.09212	0.02309	31	1.34	125.00	0
125	Glutamic (%)	0226	1.8190	0.07550	1.6212	0.09212	0.02309	31	2.15	125.00	0
125	Glutamic (%)	0619	1.8600	0.02000	1.6212	0.09212	0.02309	31	2.59	125.00	0
125	Glutamic (%)	0682	1.9940	0.00000	1.6212	0.09212	0.02309	31	4.05	125.00	0
125	Glutamic (%)	0853	2.3100	0.22000	1.6212	0.09212	0.02309	31	7.48	125.05	1
126	Glycine (%)	0353	0.50500	0.01000	0.57969	0.02752	0.01166	31	-2.71	126.00	0
126	Glycine (%)	0684	0.54450	0.00700	0.57969	0.02752	0.01166	31	-1.28	126.00	0
126	Glycine (%)	0675	0.54500	0.01000	0.57969	0.02752	0.01166	31	-1.26	126.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
126	Glycine (%)	2188	0.54500	0.01000	0.57969	0.02752	0.01166	31	-1.26	126.05	0
126	Glycine (%)	0148	0.55500	0.01000	0.57969	0.02752	0.01166	31	-0.90	126.05	0
126	Glycine (%)	0870	0.55685	0.00430	0.57969	0.02752	0.01166	31	-0.83	126.00	0
126	Glycine (%)	0354	0.55800	0.00000	0.57969	0.02752	0.01166	31	-0.79	126.00	0
126	Glycine (%)	0662	0.55805	0.01610	0.57969	0.02752	0.01166	31	-0.79	126.00	0
126	Glycine (%)	0676	0.56500	0.01000	0.57969	0.02752	0.01166	31	-0.53	126.05	0
126	Glycine (%)	0407	0.56800	0.01200	0.57969	0.02752	0.01166	31	-0.42	126.05	0
126	Glycine (%)	0227	0.57000	0.00000	0.57969	0.02752	0.01166	31	-0.35	126.00	0
126	Glycine (%)	0098	0.57200	0.00400	0.57969	0.02752	0.01166	31	-0.28	126.02	0
126	Glycine (%)	0208	0.57500	0.01000	0.57969	0.02752	0.01166	31	-0.17	126.00	0
126	Glycine (%)	0910	0.57500	0.03000	0.57969	0.02752	0.01166	31	-0.17	126.00	0
126	Glycine (%)	0571	0.57550	0.00100	0.57969	0.02752	0.01166	31	-0.15	126.00	0
126	Glycine (%)	0644	0.57850	0.00500	0.57969	0.02752	0.01166	31	-0.04	126.00	0
126	Glycine (%)	0652	0.58000	0.00000	0.57969	0.02752	0.01166	31	0.01	126.00	0
126	Glycine (%)	2196	0.58100	0.00000	0.57969	0.02752	0.01166	31	0.05	126.05	0
126	Glycine (%)	0872	0.58350	0.00500	0.57969	0.02752	0.01166	31	0.14	126.00	0
126	Glycine (%)	0504	0.58500	0.03000	0.57969	0.02752	0.01166	31	0.19	126.00	0
126	Glycine (%)	0859	0.58700	0.00600	0.57969	0.02752	0.01166	31	0.27	126.00	0
126	Glycine (%)	0968	0.58700	0.00400	0.57969	0.02752	0.01166	31	0.27	126.00	0
126	Glycine (%)	2059	0.59250	0.00500	0.57969	0.02752	0.01166	31	0.47	126.00	0
126	Glycine (%)	0868	0.59450	0.01500	0.57969	0.02752	0.01166	31	0.54	126.00	0
126	Glycine (%)	0878	0.59500	0.00400	0.57969	0.02752	0.01166	31	0.56	126.00	0
126	Glycine (%)	0918	0.60130	0.00720	0.57969	0.02752	0.01166	31	0.79	126.00	0
126	Glycine (%)	0941	0.62000	0.00600	0.57969	0.02752	0.01166	31	1.46	126.00	0
126	Glycine (%)	0226	0.62295	0.04590	0.57969	0.02752	0.01166	31	1.57	126.00	0
126	Glycine (%)	0619	0.62800	0.01400	0.57969	0.02752	0.01166	31	1.76	126.00	0
126	Glycine (%)	0853	0.65000	0.08000	0.57969	0.02752	0.01166	31	2.56	126.05	0
126	Glycine (%)	0682	0.68300	0.00000	0.57969	0.02752	0.01166	31	3.75	126.00	0
126	Glycine (%)	0626	289.78	578.43	0.57969	0.02752	0.01166	31	10510.39	126.05	2
127	Histidine (%)	0853	0.05000	0.02000	0.26516	0.01604	0.00520	31	-13.42	127.05	0
127	Histidine (%)	0675	0.23500	0.01000	0.26516	0.01604	0.00520	31	-1.88	127.00	0
127	Histidine (%)	0353	0.24000	0.00000	0.26516	0.01604	0.00520	31	-1.57	127.00	0
127	Histidine (%)	0626	0.24450	0.00700	0.26516	0.01604	0.00520	31	-1.29	127.05	0
127	Histidine (%)	0870	0.24835	0.00350	0.26516	0.01604	0.00520	31	-1.05	127.00	0
127	Histidine (%)	0684	0.25050	0.00300	0.26516	0.01604	0.00520	31	-0.91	127.00	0
127	Histidine (%)	0354	0.25250	0.00100	0.26516	0.01604	0.00520	31	-0.79	127.00	0
127	Histidine (%)	0910	0.25500	0.03000	0.26516	0.01604	0.00520	31	-0.63	127.00	0
127	Histidine (%)	2196	0.25500	0.00000	0.26516	0.01604	0.00520	31	-0.63	127.05	0
127	Histidine (%)	0878	0.25800	0.00000	0.26516	0.01604	0.00520	31	-0.45	127.00	0
127	Histidine (%)	0968	0.25800	0.00400	0.26516	0.01604	0.00520	31	-0.45	127.00	0
127	Histidine (%)	0098	0.26100	0.00800	0.26516	0.01604	0.00520	31	-0.26	127.02	0
127	Histidine (%)	2059	0.26350	0.00300	0.26516	0.01604	0.00520	31	-0.10	127.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
127	Histidine (%)	0941	0.26400	0.01000	0.26516	0.01604	0.00520	31	-0.07	127.00	0
127	Histidine (%)	0571	0.26450	0.00100	0.26516	0.01604	0.00520	31	-0.04	127.00	0
127	Histidine (%)	0226	0.26845	0.01010	0.26516	0.01604	0.00520	31	0.21	127.00	0
127	Histidine (%)	0208	0.27000	0.00000	0.26516	0.01604	0.00520	31	0.30	127.00	0
127	Histidine (%)	0227	0.27000	0.00000	0.26516	0.01604	0.00520	31	0.30	127.00	0
127	Histidine (%)	0644	0.27000	0.00200	0.26516	0.01604	0.00520	31	0.30	127.00	0
127	Histidine (%)	0148	0.27000	0.00000	0.26516	0.01604	0.00520	31	0.30	127.05	0
127	Histidine (%)	0676	0.27000	0.00000	0.26516	0.01604	0.00520	31	0.30	127.05	0
127	Histidine (%)	0918	0.27035	0.00450	0.26516	0.01604	0.00520	31	0.32	127.00	0
127	Histidine (%)	0652	0.27500	0.01000	0.26516	0.01604	0.00520	31	0.61	127.00	0
127	Histidine (%)	0868	0.27500	0.00400	0.26516	0.01604	0.00520	31	0.61	127.00	0
127	Histidine (%)	0859	0.27600	0.00400	0.26516	0.01604	0.00520	31	0.68	127.00	0
127	Histidine (%)	0872	0.27750	0.00100	0.26516	0.01604	0.00520	31	0.77	127.00	0
127	Histidine (%)	0662	0.28030	0.01120	0.26516	0.01604	0.00520	31	0.94	127.00	0
127	Histidine (%)	0407	0.28150	0.00500	0.26516	0.01604	0.00520	31	1.02	127.05	0
127	Histidine (%)	0504	0.29000	0.00000	0.26516	0.01604	0.00520	31	1.55	127.00	0
127	Histidine (%)	0619	0.34050	0.00900	0.26516	0.01604	0.00520	31	4.70	127.00	0
127	Histidine (%)	0682	0.39400	0.00000	0.26516	0.01604	0.00520	31	8.03	127.00	0
127	Histidine (%)	2188	1.9200	0.08000	0.26516	0.01604	0.00520	31	103.20	127.05	2
128	Isoleucine (%)	0853	0.21500	0.01000	0.34219	0.02508	0.00506	31	-5.07	128.05	0
128	Isoleucine (%)	0353	0.26000	0.00000	0.34219	0.02508	0.00506	31	-3.28	128.00	0
128	Isoleucine (%)	0626	0.30450	0.00100	0.34219	0.02508	0.00506	31	-1.50	128.05	0
128	Isoleucine (%)	0675	0.31000	0.00000	0.34219	0.02508	0.00506	31	-1.28	128.00	0
128	Isoleucine (%)	0682	0.31000	0.00000	0.34219	0.02508	0.00506	31	-1.28	128.00	0
128	Isoleucine (%)	0870	0.31250	0.01640	0.34219	0.02508	0.00506	31	-1.18	128.00	0
128	Isoleucine (%)	0676	0.32000	0.00000	0.34219	0.02508	0.00506	31	-0.88	128.05	0
128	Isoleucine (%)	0684	0.32550	0.01300	0.34219	0.02508	0.00506	31	-0.67	128.00	0
128	Isoleucine (%)	0354	0.33050	0.00100	0.34219	0.02508	0.00506	31	-0.47	128.00	0
128	Isoleucine (%)	0968	0.33400	0.00000	0.34219	0.02508	0.00506	31	-0.33	128.00	0
128	Isoleucine (%)	0098	0.33950	0.01300	0.34219	0.02508	0.00506	31	-0.11	128.02	0
128	Isoleucine (%)	2188	0.34000	0.00000	0.34219	0.02508	0.00506	31	-0.09	128.05	0
128	Isoleucine (%)	0859	0.34350	0.00700	0.34219	0.02508	0.00506	31	0.05	128.00	0
128	Isoleucine (%)	0878	0.34350	0.01700	0.34219	0.02508	0.00506	31	0.05	128.00	0
128	Isoleucine (%)	0918	0.34410	0.00420	0.34219	0.02508	0.00506	31	0.08	128.00	0
128	Isoleucine (%)	0227	0.34500	0.01000	0.34219	0.02508	0.00506	31	0.11	128.00	0
128	Isoleucine (%)	0619	0.34750	0.00500	0.34219	0.02508	0.00506	31	0.21	128.00	0
128	Isoleucine (%)	0644	0.34900	0.00200	0.34219	0.02508	0.00506	31	0.27	128.00	0
128	Isoleucine (%)	0208	0.35000	0.00000	0.34219	0.02508	0.00506	31	0.31	128.00	0
128	Isoleucine (%)	0652	0.35000	0.00000	0.34219	0.02508	0.00506	31	0.31	128.00	0
128	Isoleucine (%)	0868	0.35400	0.00800	0.34219	0.02508	0.00506	31	0.47	128.00	0
128	Isoleucine (%)	0662	0.35410	0.01040	0.34219	0.02508	0.00506	31	0.47	128.00	0
128	Isoleucine (%)	0910	0.35500	0.01000	0.34219	0.02508	0.00506	31	0.51	128.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
128	Isoleucine (%)	2059	0.35550	0.00300	0.34219	0.02508	0.00506	31	0.53	128.00	0
128	Isoleucine (%)	0571	0.35650	0.00100	0.34219	0.02508	0.00506	31	0.57	128.00	0
128	Isoleucine (%)	0872	0.35750	0.00100	0.34219	0.02508	0.00506	31	0.61	128.00	0
128	Isoleucine (%)	0941	0.35750	0.00100	0.34219	0.02508	0.00506	31	0.61	128.00	0
128	Isoleucine (%)	0504	0.37000	0.00000	0.34219	0.02508	0.00506	31	1.11	128.00	0
128	Isoleucine (%)	0148	0.38000	0.02000	0.34219	0.02508	0.00506	31	1.51	128.05	0
128	Isoleucine (%)	2196	0.38100	0.00000	0.34219	0.02508	0.00506	31	1.55	128.05	0
128	Isoleucine (%)	0226	0.42305	0.00290	0.34219	0.02508	0.00506	31	3.22	128.00	0
128	Isoleucine (%)	0407	0.32850	0.07700	0.34219	0.02508	0.00506	31	-0.55	128.05	1
129	Leucine (%)	0353	0.71500	0.01000	0.79342	0.03494	0.00927	31	-2.24	129.00	0
129	Leucine (%)	2188	0.74500	0.03000	0.79342	0.03494	0.00927	31	-1.39	129.05	0
129	Leucine (%)	0682	0.75200	0.00000	0.79342	0.03494	0.00927	31	-1.19	129.00	0
129	Leucine (%)	0407	0.75200	0.00400	0.79342	0.03494	0.00927	31	-1.19	129.05	0
129	Leucine (%)	0626	0.75500	0.00000	0.79342	0.03494	0.00927	31	-1.10	129.05	0
129	Leucine (%)	0870	0.75845	0.02610	0.79342	0.03494	0.00927	31	-1.00	129.00	0
129	Leucine (%)	0684	0.76250	0.02100	0.79342	0.03494	0.00927	31	-0.89	129.00	0
129	Leucine (%)	0354	0.76750	0.00300	0.79342	0.03494	0.00927	31	-0.74	129.00	0
129	Leucine (%)	0675	0.77000	0.02000	0.79342	0.03494	0.00927	31	-0.67	129.00	0
129	Leucine (%)	0676	0.77000	0.00000	0.79342	0.03494	0.00927	31	-0.67	129.05	0
129	Leucine (%)	0208	0.78500	0.01000	0.79342	0.03494	0.00927	31	-0.24	129.00	0
129	Leucine (%)	0872	0.78650	0.00300	0.79342	0.03494	0.00927	31	-0.20	129.00	0
129	Leucine (%)	0227	0.79000	0.00000	0.79342	0.03494	0.00927	31	-0.10	129.00	0
129	Leucine (%)	0652	0.79000	0.00000	0.79342	0.03494	0.00927	31	-0.10	129.00	0
129	Leucine (%)	0878	0.79350	0.00100	0.79342	0.03494	0.00927	31	0.00	129.00	0
129	Leucine (%)	0098	0.79550	0.00500	0.79342	0.03494	0.00927	31	0.06	129.02	0
129	Leucine (%)	0644	0.79800	0.00200	0.79342	0.03494	0.00927	31	0.13	129.00	0
129	Leucine (%)	0910	0.80000	0.02000	0.79342	0.03494	0.00927	31	0.19	129.00	0
129	Leucine (%)	0148	0.80000	0.02000	0.79342	0.03494	0.00927	31	0.19	129.05	0
129	Leucine (%)	2059	0.80250	0.00300	0.79342	0.03494	0.00927	31	0.26	129.00	0
129	Leucine (%)	0571	0.80300	0.00000	0.79342	0.03494	0.00927	31	0.27	129.00	0
129	Leucine (%)	2196	0.80600	0.00000	0.79342	0.03494	0.00927	31	0.36	129.05	0
129	Leucine (%)	0504	0.81000	0.00000	0.79342	0.03494	0.00927	31	0.47	129.00	0
129	Leucine (%)	0662	0.81695	0.03090	0.79342	0.03494	0.00927	31	0.67	129.00	0
129	Leucine (%)	0868	0.81700	0.01200	0.79342	0.03494	0.00927	31	0.67	129.00	0
129	Leucine (%)	0968	0.82400	0.00400	0.79342	0.03494	0.00927	31	0.88	129.00	0
129	Leucine (%)	0918	0.82470	0.00740	0.79342	0.03494	0.00927	31	0.90	129.00	0
129	Leucine (%)	0859	0.84250	0.02100	0.79342	0.03494	0.00927	31	1.40	129.00	0
129	Leucine (%)	0941	0.85600	0.00200	0.79342	0.03494	0.00927	31	1.79	129.00	0
129	Leucine (%)	0619	0.87750	0.02300	0.79342	0.03494	0.00927	31	2.41	129.00	0
129	Leucine (%)	0226	0.90650	0.00900	0.79342	0.03494	0.00927	31	3.24	129.00	0
129	Leucine (%)	0853	0.93000	0.10000	0.79342	0.03494	0.00927	31	3.91	129.05	1
130	L-Lysine (%)	0853	0.27000	0.04000	0.40288	0.02866	0.01000	34	-4.64	130.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
130	L-Lysine (%)	0227	0.33000	0.02000	0.40288	0.02866	0.01000	34	-2.54	130.00	0
130	L-Lysine (%)	0353	0.35000	0.00000	0.40288	0.02866	0.01000	34	-1.85	130.00	0
130	L-Lysine (%)	2053	0.35500	0.01000	0.40288	0.02866	0.01000	34	-1.67	130.99	0
130	L-Lysine (%)	0407	0.37600	0.00800	0.40288	0.02866	0.01000	34	-0.94	130.05	0
130	L-Lysine (%)	0684	0.37950	0.01100	0.40288	0.02866	0.01000	34	-0.82	130.00	0
130	L-Lysine (%)	0676	0.38000	0.00000	0.40288	0.02866	0.01000	34	-0.80	130.05	0
130	L-Lysine (%)	0208	0.38500	0.01000	0.40288	0.02866	0.01000	34	-0.62	130.00	0
130	L-Lysine (%)	0675	0.38500	0.01000	0.40288	0.02866	0.01000	34	-0.62	130.00	0
130	L-Lysine (%)	0098	0.38550	0.00100	0.40288	0.02866	0.01000	34	-0.61	130.02	0
130	L-Lysine (%)	0354	0.38850	0.00700	0.40288	0.02866	0.01000	34	-0.50	130.00	0
130	L-Lysine (%)	0626	0.39250	0.00700	0.40288	0.02866	0.01000	34	-0.36	130.05	0
130	L-Lysine (%)	0910	0.39500	0.03000	0.40288	0.02866	0.01000	34	-0.27	130.00	0
130	L-Lysine (%)	2188	0.39500	0.03000	0.40288	0.02866	0.01000	34	-0.27	130.05	0
130	L-Lysine (%)	0872	0.39600	0.00400	0.40288	0.02866	0.01000	34	-0.24	130.00	0
130	L-Lysine (%)	0878	0.40450	0.00100	0.40288	0.02866	0.01000	34	0.06	130.00	0
130	L-Lysine (%)	0652	0.40500	0.01000	0.40288	0.02866	0.01000	34	0.07	130.00	0
130	L-Lysine (%)	0571	0.40600	0.00400	0.40288	0.02866	0.01000	34	0.11	130.00	0
130	L-Lysine (%)	0870	0.40640	0.01360	0.40288	0.02866	0.01000	34	0.12	130.00	0
130	L-Lysine (%)	0968	0.40700	0.00000	0.40288	0.02866	0.01000	34	0.14	130.00	0
130	L-Lysine (%)	0644	0.40850	0.00500	0.40288	0.02866	0.01000	34	0.20	130.00	0
130	L-Lysine (%)	0662	0.40895	0.00210	0.40288	0.02866	0.01000	34	0.21	130.00	0
130	L-Lysine (%)	0918	0.41110	0.01540	0.40288	0.02866	0.01000	34	0.29	130.00	0
130	L-Lysine (%)	2196	0.41700	0.00000	0.40288	0.02866	0.01000	34	0.49	130.05	0
130	L-Lysine (%)	0941	0.41800	0.00600	0.40288	0.02866	0.01000	34	0.53	130.00	0
130	L-Lysine (%)	2059	0.41950	0.00500	0.40288	0.02866	0.01000	34	0.58	130.00	0
130	L-Lysine (%)	0148	0.42000	0.02000	0.40288	0.02866	0.01000	34	0.60	130.05	0
130	L-Lysine (%)	0226	0.42270	0.01580	0.40288	0.02866	0.01000	34	0.69	130.00	0
130	L-Lysine (%)	0868	0.43200	0.01400	0.40288	0.02866	0.01000	34	1.02	130.00	0
130	L-Lysine (%)	0859	0.43400	0.00000	0.40288	0.02866	0.01000	34	1.09	130.00	0
130	L-Lysine (%)	0619	0.44200	0.01000	0.40288	0.02866	0.01000	34	1.37	130.00	0
130	L-Lysine (%)	0504	0.45000	0.00000	0.40288	0.02866	0.01000	34	1.64	130.00	0
130	L-Lysine (%)	2145	0.53500	0.03000	0.40288	0.02866	0.01000	34	4.61	130.99	0
130	L-Lysine (%)	0682	0.54700	0.00000	0.40288	0.02866	0.01000	34	5.03	130.00	0
131	Methionine (%)	0684	0.01300	0.00000	0.12384	0.01569	0.00531	33	-7.07	131.00	0
131	Methionine (%)	0626	0.09700	0.00400	0.12384	0.01569	0.00531	33	-1.71	131.05	0
131	Methionine (%)	0870	0.09930	0.00060	0.12384	0.01569	0.00531	33	-1.56	131.00	0
131	Methionine (%)	0941	0.09950	0.00100	0.12384	0.01569	0.00531	33	-1.55	131.00	0
131	Methionine (%)	0675	0.11000	0.00000	0.12384	0.01569	0.00531	33	-0.88	131.00	0
131	Methionine (%)	0676	0.11000	0.00000	0.12384	0.01569	0.00531	33	-0.88	131.05	0
131	Methionine (%)	2053	0.11000	0.02000	0.12384	0.01569	0.00531	33	-0.88	131.99	0
131	Methionine (%)	0227	0.11500	0.01000	0.12384	0.01569	0.00531	33	-0.56	131.00	0
131	Methionine (%)	0652	0.11500	0.01000	0.12384	0.01569	0.00531	33	-0.56	131.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
131	Methionine (%)	0148	0.11500	0.01000	0.12384	0.01569	0.00531	33	-0.56	131.05	0
131	Methionine (%)	0619	0.11650	0.00300	0.12384	0.01569	0.00531	33	-0.47	131.00	0
131	Methionine (%)	0353	0.12000	0.02000	0.12384	0.01569	0.00531	33	-0.24	131.00	0
131	Methionine (%)	0504	0.12000	0.00000	0.12384	0.01569	0.00531	33	-0.24	131.00	0
131	Methionine (%)	0098	0.12000	0.00000	0.12384	0.01569	0.00531	33	-0.24	131.02	0
131	Methionine (%)	0354	0.12300	0.00200	0.12384	0.01569	0.00531	33	-0.05	131.00	0
131	Methionine (%)	0571	0.12350	0.00100	0.12384	0.01569	0.00531	33	-0.02	131.00	0
131	Methionine (%)	0878	0.12500	0.00600	0.12384	0.01569	0.00531	33	0.07	131.00	0
131	Methionine (%)	0910	0.12500	0.01000	0.12384	0.01569	0.00531	33	0.07	131.00	0
131	Methionine (%)	0872	0.12650	0.00300	0.12384	0.01569	0.00531	33	0.17	131.00	0
131	Methionine (%)	0662	0.12730	0.00180	0.12384	0.01569	0.00531	33	0.22	131.00	0
131	Methionine (%)	0682	0.12800	0.00000	0.12384	0.01569	0.00531	33	0.27	131.00	0
131	Methionine (%)	0968	0.12900	0.00400	0.12384	0.01569	0.00531	33	0.33	131.00	0
131	Methionine (%)	0868	0.12950	0.00100	0.12384	0.01569	0.00531	33	0.36	131.00	0
131	Methionine (%)	2059	0.13050	0.00100	0.12384	0.01569	0.00531	33	0.42	131.00	0
131	Methionine (%)	0859	0.13150	0.00100	0.12384	0.01569	0.00531	33	0.49	131.00	0
131	Methionine (%)	0644	0.13500	0.00200	0.12384	0.01569	0.00531	33	0.71	131.00	0
131	Methionine (%)	2196	0.13600	0.00000	0.12384	0.01569	0.00531	33	0.78	131.05	0
131	Methionine (%)	0226	0.13795	0.00590	0.12384	0.01569	0.00531	33	0.90	131.00	0
131	Methionine (%)	0918	0.14000	0.00000	0.12384	0.01569	0.00531	33	1.03	131.00	0
131	Methionine (%)	0208	0.14400	0.02400	0.12384	0.01569	0.00531	33	1.29	131.00	0
131	Methionine (%)	0407	0.15000	0.01400	0.12384	0.01569	0.00531	33	1.67	131.05	0
131	Methionine (%)	2145	0.15500	0.01000	0.12384	0.01569	0.00531	33	1.99	131.99	0
131	Methionine (%)	2188	0.29500	0.01000	0.12384	0.01569	0.00531	33	10.91	131.05	0
131	Methionine (%)	0853	0.00500	0.01000	0.12384	0.01569	0.00531	33		131.05	4
132	Phenylalanine (%)	0853	0.33000	0.04000	0.40026	0.03590	0.00857	31	-1.96	132.05	0
132	Phenylalanine (%)	0407	0.34900	0.01600	0.40026	0.03590	0.00857	31	-1.43	132.05	0
132	Phenylalanine (%)	0941	0.36350	0.00100	0.40026	0.03590	0.00857	31	-1.02	132.00	0
132	Phenylalanine (%)	0675	0.37500	0.01000	0.40026	0.03590	0.00857	31	-0.70	132.00	0
132	Phenylalanine (%)	0684	0.37950	0.01100	0.40026	0.03590	0.00857	31	-0.58	132.00	0
132	Phenylalanine (%)	0868	0.37950	0.00900	0.40026	0.03590	0.00857	31	-0.58	132.00	0
132	Phenylalanine (%)	0872	0.38000	0.01200	0.40026	0.03590	0.00857	31	-0.56	132.00	0
132	Phenylalanine (%)	0148	0.38000	0.00000	0.40026	0.03590	0.00857	31	-0.56	132.05	0
132	Phenylalanine (%)	0878	0.38150	0.00100	0.40026	0.03590	0.00857	31	-0.52	132.00	0
132	Phenylalanine (%)	0870	0.38180	0.00740	0.40026	0.03590	0.00857	31	-0.51	132.00	0
132	Phenylalanine (%)	0098	0.38450	0.00900	0.40026	0.03590	0.00857	31	-0.44	132.02	0
132	Phenylalanine (%)	0208	0.38500	0.01000	0.40026	0.03590	0.00857	31	-0.42	132.00	0
132	Phenylalanine (%)	0227	0.38500	0.01000	0.40026	0.03590	0.00857	31	-0.42	132.00	0
132	Phenylalanine (%)	0676	0.38500	0.01000	0.40026	0.03590	0.00857	31	-0.42	132.05	0
132	Phenylalanine (%)	0626	0.38850	0.00300	0.40026	0.03590	0.00857	31	-0.33	132.05	0
132	Phenylalanine (%)	0910	0.39000	0.00000	0.40026	0.03590	0.00857	31	-0.29	132.00	0
132	Phenylalanine (%)	0644	0.39500	0.00600	0.40026	0.03590	0.00857	31	-0.15	132.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
132	Phenylalanine (%)	0571	0.39650	0.00500	0.40026	0.03590	0.00857	31	-0.10	132.00	0
132	Phenylalanine (%)	2196	0.39700	0.00000	0.40026	0.03590	0.00857	31	-0.09	132.05	0
132	Phenylalanine (%)	0354	0.39900	0.00400	0.40026	0.03590	0.00857	31	-0.03	132.00	0
132	Phenylalanine (%)	2188	0.40000	0.00000	0.40026	0.03590	0.00857	31	-0.01	132.05	0
132	Phenylalanine (%)	2059	0.40100	0.00600	0.40026	0.03590	0.00857	31	0.02	132.00	0
132	Phenylalanine (%)	0504	0.40500	0.03000	0.40026	0.03590	0.00857	31	0.13	132.00	0
132	Phenylalanine (%)	0968	0.41850	0.00100	0.40026	0.03590	0.00857	31	0.51	132.00	0
132	Phenylalanine (%)	0918	0.44680	0.00320	0.40026	0.03590	0.00857	31	1.30	132.00	0
132	Phenylalanine (%)	0682	0.44700	0.00000	0.40026	0.03590	0.00857	31	1.30	132.00	0
132	Phenylalanine (%)	0859	0.45150	0.01100	0.40026	0.03590	0.00857	31	1.43	132.00	0
132	Phenylalanine (%)	0353	0.45500	0.01000	0.40026	0.03590	0.00857	31	1.52	132.00	0
132	Phenylalanine (%)	0662	0.46310	0.02700	0.40026	0.03590	0.00857	31	1.75	132.00	0
132	Phenylalanine (%)	0226	0.48165	0.00210	0.40026	0.03590	0.00857	31	2.27	132.00	0
132	Phenylalanine (%)	0619	0.49950	0.01100	0.40026	0.03590	0.00857	31	2.76	132.00	0
132	Phenylalanine (%)	0652	0.41500	0.07000	0.40026	0.03590	0.00857	31	0.41	132.00	1
133	Proline (%)	0870	0.66190	0.01640	0.76317	0.04670	0.01748	30	-2.17	133.00	0
133	Proline (%)	0662	0.69715	0.05510	0.76317	0.04670	0.01748	30	-1.41	133.00	0
133	Proline (%)	0354	0.70700	0.01000	0.76317	0.04670	0.01748	30	-1.20	133.00	0
133	Proline (%)	0353	0.71000	0.00000	0.76317	0.04670	0.01748	30	-1.14	133.00	0
133	Proline (%)	0684	0.72500	0.00200	0.76317	0.04670	0.01748	30	-0.82	133.00	0
133	Proline (%)	0407	0.73150	0.04300	0.76317	0.04670	0.01748	30	-0.68	133.05	0
133	Proline (%)	0918	0.73370	0.01480	0.76317	0.04670	0.01748	30	-0.63	133.00	0
133	Proline (%)	0675	0.73500	0.01000	0.76317	0.04670	0.01748	30	-0.60	133.00	0
133	Proline (%)	0910	0.73500	0.01000	0.76317	0.04670	0.01748	30	-0.60	133.00	0
133	Proline (%)	0968	0.73800	0.01000	0.76317	0.04670	0.01748	30	-0.54	133.00	0
133	Proline (%)	0676	0.74500	0.01000	0.76317	0.04670	0.01748	30	-0.39	133.05	0
133	Proline (%)	0868	0.75050	0.00500	0.76317	0.04670	0.01748	30	-0.27	133.00	0
133	Proline (%)	2059	0.75150	0.00700	0.76317	0.04670	0.01748	30	-0.25	133.00	0
133	Proline (%)	0878	0.75300	0.00800	0.76317	0.04670	0.01748	30	-0.22	133.00	0
133	Proline (%)	0227	0.75500	0.03000	0.76317	0.04670	0.01748	30	-0.17	133.00	0
133	Proline (%)	0859	0.76400	0.01800	0.76317	0.04670	0.01748	30	0.02	133.00	0
133	Proline (%)	0504	0.76500	0.01000	0.76317	0.04670	0.01748	30	0.04	133.00	0
133	Proline (%)	0148	0.77000	0.00000	0.76317	0.04670	0.01748	30	0.15	133.05	0
133	Proline (%)	0644	0.77100	0.04800	0.76317	0.04670	0.01748	30	0.17	133.00	0
133	Proline (%)	0571	0.77300	0.01200	0.76317	0.04670	0.01748	30	0.21	133.00	0
133	Proline (%)	0872	0.77800	0.01200	0.76317	0.04670	0.01748	30	0.32	133.00	0
133	Proline (%)	0208	0.78000	0.02000	0.76317	0.04670	0.01748	30	0.36	133.00	0
133	Proline (%)	0652	0.78000	0.00000	0.76317	0.04670	0.01748	30	0.36	133.00	0
133	Proline (%)	0619	0.78400	0.01600	0.76317	0.04670	0.01748	30	0.45	133.00	0
133	Proline (%)	0941	0.80350	0.01700	0.76317	0.04670	0.01748	30	0.86	133.00	0
133	Proline (%)	0226	0.87670	0.11100	0.76317	0.04670	0.01748	30	2.43	133.00	0
133	Proline (%)	2196	0.90700	0.00000	0.76317	0.04670	0.01748	30	3.08	133.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
133	Proline (%)	0682	0.91400	0.00000	0.76317	0.04670	0.01748	30	3.23	133.00	0
133	Proline (%)	0626	0.93450	0.01900	0.76317	0.04670	0.01748	30	3.67	133.05	0
133	Proline (%)	2188	1.1050	0.01000	0.76317	0.04670	0.01748	30	7.32	133.05	0
133	Proline (%)	0853	0.96000	0.14000	0.76317	0.04670	0.01748	30	4.21	133.05	1
134	Serine (%)	0504	0.42000	0.02000	0.49728	0.02788	0.01279	31	-2.77	134.00	0
134	Serine (%)	0098	0.44300	0.00800	0.49728	0.02788	0.01279	31	-1.95	134.02	0
134	Serine (%)	0662	0.46380	0.00020	0.49728	0.02788	0.01279	31	-1.20	134.00	0
134	Serine (%)	0353	0.47000	0.00000	0.49728	0.02788	0.01279	31	-0.98	134.00	0
134	Serine (%)	0675	0.47000	0.02000	0.49728	0.02788	0.01279	31	-0.98	134.00	0
134	Serine (%)	0626	0.47800	0.00200	0.49728	0.02788	0.01279	31	-0.69	134.05	0
134	Serine (%)	0910	0.48000	0.00000	0.49728	0.02788	0.01279	31	-0.62	134.00	0
134	Serine (%)	0872	0.48100	0.00800	0.49728	0.02788	0.01279	31	-0.58	134.00	0
134	Serine (%)	0644	0.48150	0.01300	0.49728	0.02788	0.01279	31	-0.57	134.00	0
134	Serine (%)	0571	0.48200	0.00000	0.49728	0.02788	0.01279	31	-0.55	134.00	0
134	Serine (%)	2059	0.48350	0.00500	0.49728	0.02788	0.01279	31	-0.49	134.00	0
134	Serine (%)	0208	0.48500	0.01000	0.49728	0.02788	0.01279	31	-0.44	134.00	0
134	Serine (%)	0652	0.48500	0.01000	0.49728	0.02788	0.01279	31	-0.44	134.00	0
134	Serine (%)	0676	0.48500	0.01000	0.49728	0.02788	0.01279	31	-0.44	134.05	0
134	Serine (%)	0684	0.48850	0.02700	0.49728	0.02788	0.01279	31	-0.32	134.00	0
134	Serine (%)	0878	0.49450	0.01100	0.49728	0.02788	0.01279	31	-0.10	134.00	0
134	Serine (%)	0227	0.50000	0.00000	0.49728	0.02788	0.01279	31	0.10	134.00	0
134	Serine (%)	0407	0.50200	0.05600	0.49728	0.02788	0.01279	31	0.17	134.05	0
134	Serine (%)	0918	0.50300	0.01740	0.49728	0.02788	0.01279	31	0.21	134.00	0
134	Serine (%)	0354	0.50450	0.00100	0.49728	0.02788	0.01279	31	0.26	134.00	0
134	Serine (%)	2188	0.50500	0.01000	0.49728	0.02788	0.01279	31	0.28	134.05	0
134	Serine (%)	0941	0.50650	0.00100	0.49728	0.02788	0.01279	31	0.33	134.00	0
134	Serine (%)	0859	0.50750	0.00900	0.49728	0.02788	0.01279	31	0.37	134.00	0
134	Serine (%)	0868	0.52150	0.00700	0.49728	0.02788	0.01279	31	0.87	134.00	0
134	Serine (%)	0870	0.52245	0.00530	0.49728	0.02788	0.01279	31	0.90	134.00	0
134	Serine (%)	2196	0.52600	0.00000	0.49728	0.02788	0.01279	31	1.03	134.05	0
134	Serine (%)	0968	0.53050	0.01500	0.49728	0.02788	0.01279	31	1.19	134.00	0
134	Serine (%)	0226	0.53075	0.06070	0.49728	0.02788	0.01279	31	1.20	134.00	0
134	Serine (%)	0148	0.54500	0.05000	0.49728	0.02788	0.01279	31	1.71	134.05	0
134	Serine (%)	0619	0.57900	0.02000	0.49728	0.02788	0.01279	31	2.93	134.00	0
134	Serine (%)	0682	0.57900	0.00000	0.49728	0.02788	0.01279	31	2.93	134.00	0
134	Serine (%)	0853	0.53500	0.09000	0.49728	0.02788	0.01279	31	1.35	134.05	1
135	Threonine (%)	0407	0.32100	0.06600	0.37816	0.02156	0.00965	32	-2.65	135.05	0
135	Threonine (%)	0853	0.32500	0.05000	0.37816	0.02156	0.00965	32	-2.47	135.05	0
135	Threonine (%)	0626	0.34200	0.02200	0.37816	0.02156	0.00965	32	-1.68	135.05	0
135	Threonine (%)	0353	0.35000	0.00000	0.37816	0.02156	0.00965	32	-1.31	135.00	0
135	Threonine (%)	0662	0.35140	0.00360	0.37816	0.02156	0.00965	32	-1.24	135.00	0
135	Threonine (%)	0675	0.35500	0.03000	0.37816	0.02156	0.00965	32	-1.07	135.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
135	Threonine (%)	0676	0.35500	0.01000	0.37816	0.02156	0.00965	32	-1.07	135.05	0
135	Threonine (%)	0148	0.36500	0.01000	0.37816	0.02156	0.00965	32	-0.61	135.05	0
135	Threonine (%)	0504	0.37000	0.00000	0.37816	0.02156	0.00965	32	-0.38	135.00	0
135	Threonine (%)	0684	0.37100	0.00800	0.37816	0.02156	0.00965	32	-0.33	135.00	0
135	Threonine (%)	0571	0.37250	0.00500	0.37816	0.02156	0.00965	32	-0.26	135.00	0
135	Threonine (%)	0910	0.37500	0.01000	0.37816	0.02156	0.00965	32	-0.15	135.00	0
135	Threonine (%)	0098	0.37700	0.00600	0.37816	0.02156	0.00965	32	-0.05	135.02	0
135	Threonine (%)	0872	0.37750	0.00500	0.37816	0.02156	0.00965	32	-0.03	135.00	0
135	Threonine (%)	2059	0.37900	0.00400	0.37816	0.02156	0.00965	32	0.04	135.00	0
135	Threonine (%)	0208	0.38000	0.00000	0.37816	0.02156	0.00965	32	0.09	135.00	0
135	Threonine (%)	0652	0.38000	0.00000	0.37816	0.02156	0.00965	32	0.09	135.00	0
135	Threonine (%)	2145	0.38000	0.00000	0.37816	0.02156	0.00965	32	0.09	135.99	0
135	Threonine (%)	0878	0.38100	0.02000	0.37816	0.02156	0.00965	32	0.13	135.00	0
135	Threonine (%)	0941	0.38150	0.00900	0.37816	0.02156	0.00965	32	0.16	135.00	0
135	Threonine (%)	0644	0.38400	0.01000	0.37816	0.02156	0.00965	32	0.27	135.00	0
135	Threonine (%)	0868	0.38900	0.00000	0.37816	0.02156	0.00965	32	0.50	135.00	0
135	Threonine (%)	0354	0.38950	0.00100	0.37816	0.02156	0.00965	32	0.53	135.00	0
135	Threonine (%)	0227	0.39000	0.00000	0.37816	0.02156	0.00965	32	0.55	135.00	0
135	Threonine (%)	0870	0.39085	0.00770	0.37816	0.02156	0.00965	32	0.59	135.00	0
135	Threonine (%)	0226	0.39340	0.02300	0.37816	0.02156	0.00965	32	0.71	135.00	0
135	Threonine (%)	2196	0.39700	0.00000	0.37816	0.02156	0.00965	32	0.87	135.05	0
135	Threonine (%)	0859	0.39800	0.00400	0.37816	0.02156	0.00965	32	0.92	135.00	0
135	Threonine (%)	0918	0.39945	0.00150	0.37816	0.02156	0.00965	32	0.99	135.00	0
135	Threonine (%)	0968	0.41250	0.00100	0.37816	0.02156	0.00965	32	1.59	135.00	0
135	Threonine (%)	0619	0.42900	0.00200	0.37816	0.02156	0.00965	32	2.36	135.00	0
135	Threonine (%)	0682	0.44200	0.00000	0.37816	0.02156	0.00965	32	2.96	135.00	0
135	Threonine (%)	2188	0.37000	0.12000	0.37816	0.02156	0.00965	32	-0.38	135.05	1
136	Tryptophan (%)	0504	0.05500	0.01000	0.08779	0.01226	0.00449	18	-2.67	136.99	0
136	Tryptophan (%)	0684	0.07300	0.00600	0.08779	0.01226	0.00449	18	-1.21	136.00	0
136	Tryptophan (%)	0870	0.07590	0.00520	0.08779	0.01226	0.00449	18	-0.97	136.00	0
136	Tryptophan (%)	0208	0.07750	0.00700	0.08779	0.01226	0.00449	18	-0.84	136.01	0
136	Tryptophan (%)	0682	0.07900	0.00000	0.08779	0.01226	0.00449	18	-0.72	136.00	0
136	Tryptophan (%)	0910	0.08000	0.02000	0.08779	0.01226	0.00449	18	-0.63	136.00	0
136	Tryptophan (%)	0868	0.08050	0.00500	0.08779	0.01226	0.00449	18	-0.59	136.01	0
136	Tryptophan (%)	0872	0.08800	0.00200	0.08779	0.01226	0.00449	18	0.02	136.03	0
136	Tryptophan (%)	0878	0.08850	0.00300	0.08779	0.01226	0.00449	18	0.06	136.01	0
136	Tryptophan (%)	2059	0.09000	0.00000	0.08779	0.01226	0.00449	18	0.18	136.03	0
136	Tryptophan (%)	0098	0.09050	0.00300	0.08779	0.01226	0.00449	18	0.22	136.02	0
136	Tryptophan (%)	0859	0.09200	0.00200	0.08779	0.01226	0.00449	18	0.34	136.03	0
136	Tryptophan (%)	0407	0.09350	0.00500	0.08779	0.01226	0.00449	18	0.47	136.05	0
136	Tryptophan (%)	0619	0.09500	0.00000	0.08779	0.01226	0.00449	18	0.59	136.00	0
136	Tryptophan (%)	0571	0.09500	0.00000	0.08779	0.01226	0.00449	18	0.59	136.03	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
136	Tryptophan (%)	0227	0.10000	0.00000	0.08779	0.01226	0.00449	18	1.00	136.00	0
136	Tryptophan (%)	0662	0.10755	0.01270	0.08779	0.01226	0.00449	18	1.61	136.00	0
136	Tryptophan (%)	0353	0.11000	0.00000	0.08779	0.01226	0.00449	18	1.81	136.03	0
136	Tryptophan (%)	0918	0.13500	0.03000	0.08779	0.01226	0.00449	18	3.85	136.00	1
137	Tyrosine (%)	0910	0.08500	0.01000	0.30977	0.05400	0.01288	24	-4.16	137.00	0
137	Tyrosine (%)	0682	0.22700	0.00000	0.30977	0.05400	0.01288	24	-1.53	137.00	0
137	Tyrosine (%)	0626	0.25150	0.00500	0.30977	0.05400	0.01288	24	-1.08	137.05	0
137	Tyrosine (%)	0675	0.26000	0.02000	0.30977	0.05400	0.01288	24	-0.92	137.00	0
137	Tyrosine (%)	0208	0.27000	0.00000	0.30977	0.05400	0.01288	24	-0.74	137.00	0
137	Tyrosine (%)	0227	0.27500	0.01000	0.30977	0.05400	0.01288	24	-0.64	137.00	0
137	Tyrosine (%)	0354	0.27950	0.00500	0.30977	0.05400	0.01288	24	-0.56	137.00	0
137	Tyrosine (%)	0098	0.28550	0.00900	0.30977	0.05400	0.01288	24	-0.45	137.02	0
137	Tyrosine (%)	0504	0.29500	0.03000	0.30977	0.05400	0.01288	24	-0.27	137.00	0
137	Tyrosine (%)	0684	0.29750	0.01100	0.30977	0.05400	0.01288	24	-0.23	137.00	0
137	Tyrosine (%)	0676	0.30500	0.01000	0.30977	0.05400	0.01288	24	-0.09	137.05	0
137	Tyrosine (%)	0968	0.30700	0.00200	0.30977	0.05400	0.01288	24	-0.05	137.00	0
137	Tyrosine (%)	2196	0.31100	0.00000	0.30977	0.05400	0.01288	24	0.02	137.05	0
137	Tyrosine (%)	0918	0.31235	0.00410	0.30977	0.05400	0.01288	24	0.05	137.00	0
137	Tyrosine (%)	0644	0.31550	0.00500	0.30977	0.05400	0.01288	24	0.11	137.00	0
137	Tyrosine (%)	0353	0.32000	0.00000	0.30977	0.05400	0.01288	24	0.19	137.00	0
137	Tyrosine (%)	2059	0.32000	0.00400	0.30977	0.05400	0.01288	24	0.19	137.00	0
137	Tyrosine (%)	0872	0.32350	0.03300	0.30977	0.05400	0.01288	24	0.25	137.00	0
137	Tyrosine (%)	0407	0.33100	0.03200	0.30977	0.05400	0.01288	24	0.39	137.05	0
137	Tyrosine (%)	0226	0.36685	0.01110	0.30977	0.05400	0.01288	24	1.06	137.00	0
137	Tyrosine (%)	0148	0.38500	0.01000	0.30977	0.05400	0.01288	24	1.39	137.05	0
137	Tyrosine (%)	0853	0.38500	0.05000	0.30977	0.05400	0.01288	24	1.39	137.05	0
137	Tyrosine (%)	2188	0.39000	0.00000	0.30977	0.05400	0.01288	24	1.49	137.05	0
137	Tyrosine (%)	0662	0.45935	0.04790	0.30977	0.05400	0.01288	24	2.77	137.99	0
137	Tyrosine (%)	0941	0.28200	0.08000	0.30977	0.05400	0.01288	24	-0.51	137.00	1
137	Tyrosine (%)	0870	< 0.1		0.30977	0.05400	0.01288	24		137.00	5
138	Valine (%)	0353	0.37000	0.00000	0.48864	0.02832	0.00897	31	-4.19	138.00	0
138	Valine (%)	0853	0.41000	0.02000	0.48864	0.02832	0.00897	31	-2.78	138.05	0
138	Valine (%)	0675	0.45500	0.01000	0.48864	0.02832	0.00897	31	-1.19	138.00	0
138	Valine (%)	0870	0.45580	0.00020	0.48864	0.02832	0.00897	31	-1.16	138.00	0
138	Valine (%)	0626	0.45600	0.00600	0.48864	0.02832	0.00897	31	-1.15	138.05	0
138	Valine (%)	0684	0.45750	0.01300	0.48864	0.02832	0.00897	31	-1.10	138.00	0
138	Valine (%)	2196	0.45900	0.00000	0.48864	0.02832	0.00897	31	-1.05	138.05	0
138	Valine (%)	2188	0.46000	0.02000	0.48864	0.02832	0.00897	31	-1.01	138.05	0
138	Valine (%)	0662	0.47475	0.00830	0.48864	0.02832	0.00897	31	-0.49	138.00	0
138	Valine (%)	0652	0.47500	0.01000	0.48864	0.02832	0.00897	31	-0.48	138.00	0
138	Valine (%)	0227	0.48500	0.01000	0.48864	0.02832	0.00897	31	-0.13	138.00	0
138	Valine (%)	0676	0.48500	0.01000	0.48864	0.02832	0.00897	31	-0.13	138.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
138	Valine (%)	0872	0.48900	0.00400	0.48864	0.02832	0.00897	31	0.01	138.00	0
138	Valine (%)	0619	0.49100	0.01600	0.48864	0.02832	0.00897	31	0.08	138.00	0
138	Valine (%)	0354	0.49150	0.00700	0.48864	0.02832	0.00897	31	0.10	138.00	0
138	Valine (%)	0098	0.49300	0.01400	0.48864	0.02832	0.00897	31	0.15	138.02	0
138	Valine (%)	0504	0.49500	0.01000	0.48864	0.02832	0.00897	31	0.22	138.00	0
138	Valine (%)	0910	0.49500	0.01000	0.48864	0.02832	0.00897	31	0.22	138.00	0
138	Valine (%)	2059	0.49550	0.00500	0.48864	0.02832	0.00897	31	0.24	138.00	0
138	Valine (%)	0571	0.49600	0.00400	0.48864	0.02832	0.00897	31	0.26	138.00	0
138	Valine (%)	0968	0.49650	0.00500	0.48864	0.02832	0.00897	31	0.28	138.00	0
138	Valine (%)	0878	0.49750	0.00900	0.48864	0.02832	0.00897	31	0.31	138.00	0
138	Valine (%)	0644	0.50000	0.00200	0.48864	0.02832	0.00897	31	0.40	138.00	0
138	Valine (%)	0859	0.50500	0.01000	0.48864	0.02832	0.00897	31	0.58	138.00	0
138	Valine (%)	0148	0.50500	0.01000	0.48864	0.02832	0.00897	31	0.58	138.05	0
138	Valine (%)	0868	0.51450	0.04500	0.48864	0.02832	0.00897	31	0.91	138.00	0
138	Valine (%)	0208	0.51500	0.01000	0.48864	0.02832	0.00897	31	0.93	138.00	0
138	Valine (%)	0941	0.51950	0.00500	0.48864	0.02832	0.00897	31	1.09	138.00	0
138	Valine (%)	0918	0.54165	0.00370	0.48864	0.02832	0.00897	31	1.87	138.00	0
138	Valine (%)	0226	0.61030	0.00080	0.48864	0.02832	0.00897	31	4.30	138.00	0
138	Valine (%)	0682	0.65000	0.00000	0.48864	0.02832	0.00897	31	5.70	138.00	0
138	Valine (%)	0407	0.45700	0.08400	0.48864	0.02832	0.00897	31	-1.12	138.05	1
139	Taurine (%)	0407	0.00750	0.00100				3		139.05	0
139	Taurine (%)	0682	0.04600	0.00000				3		139.00	0
139	Taurine (%)	0504	0.15000	0.02000				3		139.00	0
139	Taurine (%)	0098	< 0.01					3		139.02	5
160	Fructose (%)	0638	0.40000	0.00000	0.51300	0.10354	0.03000	5		160.99	0
160	Fructose (%)	0619	0.42000	0.00000	0.51300	0.10354	0.03000	5		160.99	0
160	Fructose (%)	0148	0.53500	0.05000	0.51300	0.10354	0.03000	5		160.99	0
160	Fructose (%)	0227	0.56000	0.00000	0.51300	0.10354	0.03000	5		160.99	0
160	Fructose (%)	0861	0.65000	0.10000	0.51300	0.10354	0.03000	5		160.99	0
162	Glucose (%)	0619	0.16350	0.00100	0.25088	0.06475	0.00525	4		162.99	0
162	Glucose (%)	0227	0.24000	0.02000	0.25088	0.06475	0.00525	4		162.99	0
162	Glucose (%)	0638	0.30000	0.00000	0.25088	0.06475	0.00525	4		162.99	0
162	Glucose (%)	0861	0.30000	0.00000	0.25088	0.06475	0.00525	4		162.99	0
162	Glucose (%)	0148	0.31000	0.12000	0.25088	0.06475	0.00525	4		162.99	1
163	Lactose (%)	0148	0.00000	0.00000				0		163.99	4
163	Lactose (%)	0619	0.00000	0.00000				0		163.99	4
163	Lactose (%)	0861	< 0.1					0		163.99	5
163	Lactose (%)	0227	< 0.15					0		163.99	5
164	Maltose (%)	0227	0.34000	0.18000				1		164.99	0
164	Maltose (%)	0148	0.00000	0.00000				1		164.99	4
164	Maltose (%)	0861	< 0.1					1		164.99	5
165	Sucrose (%)	0861	0.90000	0.00000				2		165.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
165	Sucrose (%)	0619	0.92150	0.00100				2		165.99	0
165	Sucrose (%)	0227	0.90000	0.08000				2		165.99	1
165	Sucrose (%)	0148	0.75500	0.01000				2		165.99	2
166	Raffinose (%)	0227	0.14500	0.01000				2		166.99	0
166	Raffinose (%)	0619	0.17100	0.00200				2		166.99	0
167	Stachyose (%)	0227	0.20000	0.00000				2		167.99	0
167	Stachyose (%)	0619	0.26400	0.00200				2		167.99	0
351	Chlortetracycline (mg/kg (ppm))	0939	619.31	0.00000	865.05	142.61	43.802	14	-1.72	351.05	0
351	Chlortetracycline (mg/kg (ppm))	0009	722.24	7.3500	865.05	142.61	43.802	14	-1.00	351.03	0
351	Chlortetracycline (mg/kg (ppm))	0038	744.23	27.650	865.05	142.61	43.802	14	-0.85	351.03	0
351	Chlortetracycline (mg/kg (ppm))	2053	750.00	20.000	865.05	142.61	43.802	14	-0.81	351.04	0
351	Chlortetracycline (mg/kg (ppm))	2135	789.89	16.185	865.05	142.61	43.802	14	-0.53	351.03	0
351	Chlortetracycline (mg/kg (ppm))	0910	798.50	83.000	865.05	142.61	43.802	14	-0.47	351.03	0
351	Chlortetracycline (mg/kg (ppm))	0013	855.79	41.110	865.05	142.61	43.802	14	-0.06	351.03	0
351	Chlortetracycline (mg/kg (ppm))	0010	879.45	18.100	865.05	142.61	43.802	14	0.10	351.03	0
351	Chlortetracycline (mg/kg (ppm))	0036	931.90	16.800	865.05	142.61	43.802	14	0.47	351.03	0
351	Chlortetracycline (mg/kg (ppm))	0032	945.00	122.00	865.05	142.61	43.802	14	0.56	351.03	0
351	Chlortetracycline (mg/kg (ppm))	0964	959.55	39.198	865.05	142.61	43.802	14	0.66	351.00	0
351	Chlortetracycline (mg/kg (ppm))	0028	991.25	78.700	865.05	142.61	43.802	14	0.88	351.05	0
351	Chlortetracycline (mg/kg (ppm))	0218	1,038.3	10.138	865.05	142.61	43.802	14	1.21	351.00	0
351	Chlortetracycline (mg/kg (ppm))	0875	1,053.5	133.00	865.05	142.61	43.802	14	1.32	351.03	0
361	Lasalocid Sodium (mg/kg (ppm))	0875	2.0800	0.48000	2.6107	0.39868	0.22516	8	-1.33	361.05	0
361	Lasalocid Sodium (mg/kg (ppm))	2055	2.2300	0.34000	2.6107	0.39868	0.22516	8	-0.95	361.05	0
361	Lasalocid Sodium (mg/kg (ppm))	0038	2.5160	0.15200	2.6107	0.39868	0.22516	8	-0.24	361.03	0
361	Lasalocid Sodium (mg/kg (ppm))	0964	2.5761	0.07930	2.6107	0.39868	0.22516	8	-0.09	361.02	0
361	Lasalocid Sodium (mg/kg (ppm))	0036	2.7000	0.00000	2.6107	0.39868	0.22516	8	0.22	361.02	0
361	Lasalocid Sodium (mg/kg (ppm))	0610	2.7750	0.13000	2.6107	0.39868	0.22516	8	0.41	361.03	0
361	Lasalocid Sodium (mg/kg (ppm))	2113	2.8000	0.40000	2.6107	0.39868	0.22516	8	0.47	361.05	0
361	Lasalocid Sodium (mg/kg (ppm))	0553	4.6000	0.22000	2.6107	0.39868	0.22516	8	4.99	361.05	0
365	Monensin (mg/kg (ppm))	2127	11.075	0.07000	31.071	2.6355	1.1190	25	-7.59	365.05	0
365	Monensin (mg/kg (ppm))	2055	26.110	2.9000	31.071	2.6355	1.1190	25	-1.88	365.05	0
365	Monensin (mg/kg (ppm))	2113	28.500	1.0000	31.071	2.6355	1.1190	25	-0.98	365.05	0
365	Monensin (mg/kg (ppm))	0610	28.650	0.10000	31.071	2.6355	1.1190	25	-0.92	365.02	0
365	Monensin (mg/kg (ppm))	0610	28.950	0.36000	31.071	2.6355	1.1190	25	-0.80	365.05	0
365	Monensin (mg/kg (ppm))	0001	29.350	1.9000	31.071	2.6355	1.1190	25	-0.65	365.05	0
365	Monensin (mg/kg (ppm))	0032	29.900	0.20000	31.071	2.6355	1.1190	25	-0.44	365.02	0
365	Monensin (mg/kg (ppm))	0904	29.955	0.63000	31.071	2.6355	1.1190	25	-0.42	365.02	0
365	Monensin (mg/kg (ppm))	0013	29.980	0.00000	31.071	2.6355	1.1190	25	-0.41	365.04	0
365	Monensin (mg/kg (ppm))	0905	30.000	0.00000	31.071	2.6355	1.1190	25	-0.41	365.02	0
365	Monensin (mg/kg (ppm))	0010	30.350	3.3000	31.071	2.6355	1.1190	25	-0.27	365.03	0
365	Monensin (mg/kg (ppm))	0910	30.400	0.60000	31.071	2.6355	1.1190	25	-0.25	365.02	0
365	Monensin (mg/kg (ppm))	0668	30.900	0.40000	31.071	2.6355	1.1190	25	-0.06	365.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
365	Monensin (mg/kg (ppm))	0036	30.950	0.50000	31.071	2.6355	1.1190	25	-0.05	365.03	0
365	Monensin (mg/kg (ppm))	0668	31.055	0.39000	31.071	2.6355	1.1190	25	-0.01	365.99	0
365	Monensin (mg/kg (ppm))	0218	31.058	0.39700	31.071	2.6355	1.1190	25	-0.01	365.03	0
365	Monensin (mg/kg (ppm))	0009	31.090	0.94000	31.071	2.6355	1.1190	25	0.01	365.03	0
365	Monensin (mg/kg (ppm))	2053	32.500	0.60000	31.071	2.6355	1.1190	25	0.54	365.04	0
365	Monensin (mg/kg (ppm))	0964	32.860	1.1976	31.071	2.6355	1.1190	25	0.68	365.04	0
365	Monensin (mg/kg (ppm))	0875	33.200	0.40000	31.071	2.6355	1.1190	25	0.81	365.05	0
365	Monensin (mg/kg (ppm))	0027	33.970	1.6400	31.071	2.6355	1.1190	25	1.10	365.00	0
365	Monensin (mg/kg (ppm))	0028	34.000	5.0000	31.071	2.6355	1.1190	25	1.11	365.03	0
365	Monensin (mg/kg (ppm))	0171	34.875	1.5500	31.071	2.6355	1.1190	25	1.44	365.99	0
365	Monensin (mg/kg (ppm))	0033	39.300	2.2000	31.071	2.6355	1.1190	25	3.12	365.99	0
365	Monensin (mg/kg (ppm))	0553	40.150	1.7000	31.071	2.6355	1.1190	25	3.44	365.05	0
365	Monensin (mg/kg (ppm))	0038	36.440	9.7600	31.071	2.6355	1.1190	25	2.04	365.05	1
373	Oxytetracycline (mg/kg (ppm))	0028	483.60	6.4000	832.50	188.00	32.589	9	-1.86	373.06	0
373	Oxytetracycline (mg/kg (ppm))	0939	604.38	0.00000	832.50	188.00	32.589	9	-1.21	373.06	0
373	Oxytetracycline (mg/kg (ppm))	0160	787.50	29.000	832.50	188.00	32.589	9	-0.24	373.99	0
373	Oxytetracycline (mg/kg (ppm))	0033	798.00	14.000	832.50	188.00	32.589	9	-0.18	373.04	0
373	Oxytetracycline (mg/kg (ppm))	0218	893.90	16.200	832.50	188.00	32.589	9	0.33	373.03	0
373	Oxytetracycline (mg/kg (ppm))	2053	897.50	51.000	832.50	188.00	32.589	9	0.35	373.05	0
373	Oxytetracycline (mg/kg (ppm))	0036	950.75	18.700	832.50	188.00	32.589	9	0.63	373.03	0
373	Oxytetracycline (mg/kg (ppm))	0032	962.50	101.00	832.50	188.00	32.589	9	0.69	373.03	0
373	Oxytetracycline (mg/kg (ppm))	0038	1,047.5	57.000	832.50	188.00	32.589	9	1.14	373.04	0
388	Tylosin (mg/kg (ppm))	2145	52.420	0.86000	79.731	34.006	9.2857	10	-0.80	388.99	0
388	Tylosin (mg/kg (ppm))	0028	55.000	13.800	79.731	34.006	9.2857	10	-0.73	388.05	0
388	Tylosin (mg/kg (ppm))	0875	55.400	0.20000	79.731	34.006	9.2857	10	-0.72	388.05	0
388	Tylosin (mg/kg (ppm))	0910	57.400	11.600	79.731	34.006	9.2857	10	-0.66	388.03	0
388	Tylosin (mg/kg (ppm))	2144	62.600	0.80000	79.731	34.006	9.2857	10	-0.50	388.05	0
388	Tylosin (mg/kg (ppm))	0032	64.000	6.8000	79.731	34.006	9.2857	10	-0.46	388.03	0
388	Tylosin (mg/kg (ppm))	0218	93.526	5.6165	79.731	34.006	9.2857	10	0.41	388.00	0
388	Tylosin (mg/kg (ppm))	0010	106.65	18.100	79.731	34.006	9.2857	10	0.79	388.05	0
388	Tylosin (mg/kg (ppm))	0964	119.57	2.5904	79.731	34.006	9.2857	10	1.17	388.00	0
388	Tylosin (mg/kg (ppm))	0904	148.80	32.490	79.731	34.006	9.2857	10	2.03	388.03	0
400	Water Activity (Units)	0407	0.54540	0.00160	0.55541	0.00863	0.00309	9	-1.16	400.01	0
400	Water Activity (Units)	0843	0.54900	0.00200	0.55541	0.00863	0.00309	9	-0.74	400.99	0
400	Water Activity (Units)	0843	0.55115	0.00030	0.55541	0.00863	0.00309	9	-0.49	400.01	0
400	Water Activity (Units)	0942	0.55200	0.00400	0.55541	0.00863	0.00309	9	-0.39	400.01	0
400	Water Activity (Units)	0589	0.55250	0.00700	0.55541	0.00863	0.00309	9	-0.34	400.01	0
400	Water Activity (Units)	2009	0.55550	0.00100	0.55541	0.00863	0.00309	9	0.01	400.01	0
400	Water Activity (Units)	2181	0.55925	0.00170	0.55541	0.00863	0.00309	9	0.45	400.01	0
400	Water Activity (Units)	2073	0.56550	0.00900	0.55541	0.00863	0.00309	9	1.17	400.99	0
400	Water Activity (Units)	2109	0.57480	0.00120	0.55541	0.00863	0.00309	9	2.25	400.01	0
412	Starch, Dietary (%)	0353	8.6950	0.09000				1		412.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
516	Arsenic, Total (mg / kg (ppm))	0186	0.54300	0.02400	0.66153	0.12727	0.02244	10	-0.93	516.52	0
516	Arsenic, Total (mg / kg (ppm))	0910	0.56000	0.04000	0.66153	0.12727	0.02244	10	-0.80	516.52	0
516	Arsenic, Total (mg / kg (ppm))	0227	0.56650	0.00700	0.66153	0.12727	0.02244	10	-0.75	516.53	0
516	Arsenic, Total (mg / kg (ppm))	0045	0.57350	0.02900	0.66153	0.12727	0.02244	10	-0.69	516.00	0
516	Arsenic, Total (mg / kg (ppm))	0218	0.62000	0.00000	0.66153	0.12727	0.02244	10	-0.33	516.53	0
516	Arsenic, Total (mg / kg (ppm))	0553	0.64150	0.02900	0.66153	0.12727	0.02244	10	-0.16	516.53	0
516	Arsenic, Total (mg / kg (ppm))	2146	0.67000	0.04000	0.66153	0.12727	0.02244	10	0.07	516.52	0
516	Arsenic, Total (mg / kg (ppm))	0918	0.79280	0.01540	0.66153	0.12727	0.02244	10	1.03	516.53	0
516	Arsenic, Total (mg / kg (ppm))	2012	0.82300	0.03200	0.66153	0.12727	0.02244	10	1.27	516.53	0
516	Arsenic, Total (mg / kg (ppm))	0098	0.82500	0.00800	0.66153	0.12727	0.02244	10	1.28	516.53	0
516	Arsenic, Total (mg / kg (ppm))	0407	1.1449	0.68620	0.66153	0.12727	0.02244	10	3.80	516.43	1
516	Arsenic, Total (mg / kg (ppm))	0218	< 5		0.66153	0.12727	0.02244	10		516.43	5
518	Cadmium (mg / kg (ppm))	0171	0.21750	0.02300	0.28163	0.02574	0.01468	12	-2.49	518.41	0
518	Cadmium (mg / kg (ppm))	2055	0.22850	0.01500	0.28163	0.02574	0.01468	12	-2.06	518.53	0
518	Cadmium (mg / kg (ppm))	0218	0.26500	0.01000	0.28163	0.02574	0.01468	12	-0.65	518.53	0
518	Cadmium (mg / kg (ppm))	0910	0.27000	0.02000	0.28163	0.02574	0.01468	12	-0.45	518.52	0
518	Cadmium (mg / kg (ppm))	0968	0.28150	0.00100	0.28163	0.02574	0.01468	12	-0.01	518.43	0
518	Cadmium (mg / kg (ppm))	0918	0.28250	0.03020	0.28163	0.02574	0.01468	12	0.03	518.53	0
518	Cadmium (mg / kg (ppm))	0553	0.28950	0.00300	0.28163	0.02574	0.01468	12	0.31	518.53	0
518	Cadmium (mg / kg (ppm))	0186	0.29150	0.01700	0.28163	0.02574	0.01468	12	0.38	518.52	0
518	Cadmium (mg / kg (ppm))	2146	0.29500	0.01000	0.28163	0.02574	0.01468	12	0.52	518.52	0
518	Cadmium (mg / kg (ppm))	0227	0.29800	0.02200	0.28163	0.02574	0.01468	12	0.64	518.53	0
518	Cadmium (mg / kg (ppm))	0098	0.30650	0.00900	0.28163	0.02574	0.01468	12	0.97	518.53	0
518	Cadmium (mg / kg (ppm))	2012	0.31400	0.01600	0.28163	0.02574	0.01468	12	1.26	518.53	0
518	Cadmium (mg / kg (ppm))	0905	< 0.3		0.28163	0.02574	0.01468	12		518.31	5
518	Cadmium (mg / kg (ppm))	0218	< 2		0.28163	0.02574	0.01468	12		518.43	5
520	Chromium (mg / kg (ppm))	0186	10.850	0.90000	22.587	8.5684	0.94703	10	-1.37	520.52	0
520	Chromium (mg / kg (ppm))	0171	13.542	0.86000	22.587	8.5684	0.94703	10	-1.06	520.41	0
520	Chromium (mg / kg (ppm))	0918	16.218	2.2683	22.587	8.5684	0.94703	10	-0.74	520.53	0
520	Chromium (mg / kg (ppm))	0968	17.135	0.13000	22.587	8.5684	0.94703	10	-0.64	520.43	0
520	Chromium (mg / kg (ppm))	0693	23.446	0.61200	22.587	8.5684	0.94703	10	0.10	520.42	0
520	Chromium (mg / kg (ppm))	0045	26.000	4.4000	22.587	8.5684	0.94703	10	0.40	520.42	0
520	Chromium (mg / kg (ppm))	0218	28.000	0.00000	22.587	8.5684	0.94703	10	0.63	520.43	0
520	Chromium (mg / kg (ppm))	0510	28.455	0.07000	22.587	8.5684	0.94703	10	0.68	520.43	0
520	Chromium (mg / kg (ppm))	0218	30.000	0.00000	22.587	8.5684	0.94703	10	0.87	520.53	0
520	Chromium (mg / kg (ppm))	0032	32.225	0.23000	22.587	8.5684	0.94703	10	1.12	520.53	0
520	Chromium (mg / kg (ppm))	2055	15.778	13.048	22.587	8.5684	0.94703	10	-0.79	520.53	1
526	Lead (mg / kg (ppm))	2152	0.25000	0.02000	0.30769	0.04266	0.03765	11	-1.35	526.41	0
526	Lead (mg / kg (ppm))	0218	0.25500	0.05000	0.30769	0.04266	0.03765	11	-1.24	526.53	0
526	Lead (mg / kg (ppm))	0098	0.27700	0.00800	0.30769	0.04266	0.03765	11	-0.72	526.53	0
526	Lead (mg / kg (ppm))	0910	0.29500	0.05000	0.30769	0.04266	0.03765	11	-0.30	526.52	0
526	Lead (mg / kg (ppm))	0553	0.30000	0.01000	0.30769	0.04266	0.03765	11	-0.18	526.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
526	Lead (mg / kg (ppm))	0186	0.31100	0.03000	0.30769	0.04266	0.03765	11	0.08	526.52	0
526	Lead (mg / kg (ppm))	0227	0.31150	0.03900	0.30769	0.04266	0.03765	11	0.09	526.53	0
526	Lead (mg / kg (ppm))	2012	0.33250	0.05100	0.30769	0.04266	0.03765	11	0.58	526.53	0
526	Lead (mg / kg (ppm))	0918	0.34045	0.08110	0.30769	0.04266	0.03765	11	0.77	526.53	0
526	Lead (mg / kg (ppm))	0032	0.34050	0.01500	0.30769	0.04266	0.03765	11	0.77	526.53	0
526	Lead (mg / kg (ppm))	2146	0.50000	0.06000	0.30769	0.04266	0.03765	11	4.51	526.52	0
526	Lead (mg / kg (ppm))	0171	0.35200	0.30400	0.30769	0.04266	0.03765	11	1.04	526.41	1
526	Lead (mg / kg (ppm))	0905	< 3		0.30769	0.04266	0.03765	11		526.31	5
526	Lead (mg / kg (ppm))	0218	< 5		0.30769	0.04266	0.03765	11		526.43	5
529	Mercury (µg / kg (ppb))	2012	5.0370	0.81800				1		529.99	0
539	Nickel (mg / kg (ppm))	0186	6.0550	0.53000	10.919	4.7350	0.34754	5		539.52	0
539	Nickel (mg / kg (ppm))	0186	7.8500	0.46000	10.919	4.7350	0.34754	5		539.42	0
539	Nickel (mg / kg (ppm))	0918	8.6899	0.74770	10.919	4.7350	0.34754	5		539.53	0
539	Nickel (mg / kg (ppm))	0218	16.000	0.00000	10.919	4.7350	0.34754	5		539.43	0
539	Nickel (mg / kg (ppm))	0218	16.000	0.00000	10.919	4.7350	0.34754	5		539.53	0
702	Butyric Acid (4:0) (%)	0619	0.00000	0.00000				0		702.00	4
704	Caproic Acid (6:0) (%)	0619	0.00100	0.00000				1		704.00	0
706	Caprylic acid (8:0) (%) (w/w)	0619	0.00000	0.00000				0		706.01	4
708	Capric acid (10:0) (%) (w/w)	0619	0.00000	0.00000				0		708.01	4
710	Lauric Acid (12:0) (%) (w/w)	0226	0.00170	0.00000				2		710.99	0
710	Lauric Acid (12:0) (%) (w/w)	0676	0.00300	0.00000				2		710.99	0
710	Lauric Acid (12:0) (%) (w/w)	0619	0.00000	0.00000				2		710.01	4
710	Lauric Acid (12:0) (%) (w/w)	0164	< 0.02					2		710.99	5
714	Myristic Acid (14:0) (%) (w/w)	0226	0.00960	0.00020				3		714.99	0
714	Myristic Acid (14:0) (%) (w/w)	0164	0.02000	0.00000				3		714.99	0
714	Myristic Acid (14:0) (%) (w/w)	0619	0.03400	0.00200				3		714.01	0
716	Palmitic Acid (16:0) (%) (w/w)	0226	0.46045	0.01470				3		716.99	0
716	Palmitic Acid (16:0) (%) (w/w)	0619	0.53650	0.06500				3		716.01	0
716	Palmitic Acid (16:0) (%) (w/w)	0164	0.55500	0.01000				3		716.99	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0226	0.01695	0.00030	0.02411	0.00800	0.00083	4		718.99	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0676	0.01750	0.00100	0.02411	0.00800	0.00083	4		718.99	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0164	0.03000	0.00000	0.02411	0.00800	0.00083	4		718.99	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0619	0.03200	0.00200	0.02411	0.00800	0.00083	4		718.01	0
722	Stearic Acid (18:0) (%) (w/w)	0619	0.09800	0.00600				2		722.01	0
722	Stearic Acid (18:0) (%) (w/w)	0226	0.10975	0.00410				2		722.99	0
724	Oleic Acid (9c-18:1) (%) (w/w)	0226	0.59480	0.01660				3		724.99	0
724	Oleic Acid (9c-18:1) (%) (w/w)	0164	0.62500	0.01000				3		724.99	0
724	Oleic Acid (9c-18:1) (%) (w/w)	0619	0.65100	0.06000				3		724.01	0
726	Linoleic Acid (9c,12c-18:2) (%) (w/w)	0148	0.81200	0.02200	0.93414	0.11548	0.02143	4		726.02	0
726	Linoleic Acid (9c,12c-18:2) (%) (w/w)	0164	0.86500	0.03000	0.93414	0.11548	0.02143	4		726.99	0
726	Linoleic Acid (9c,12c-18:2) (%) (w/w)	0226	0.99805	0.02870	0.93414	0.11548	0.02143	4		726.99	0
726	Linoleic Acid (9c,12c-18:2) (%) (w/w)	0676	1.0615	0.00500	0.93414	0.11548	0.02143	4		726.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0619	1.1950	0.13000	0.93414	0.11548	0.02143	4	726.01	1	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v	0164	0.08000	0.00000	0.09286	0.01277	0.00223	4	728.99	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v	0619	0.08650	0.00300	0.09286	0.01277	0.00223	4	728.01	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v	0226	0.09545	0.00290	0.09286	0.01277	0.00223	4	728.99	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v	0676	0.10950	0.00300	0.09286	0.01277	0.00223	4	728.99	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v	0171	3.2500	0.10000	0.09286	0.01277	0.00223	4	728.99	2	
730	Arachidic Acid (20:0) (% (w/w))	0226	0.01065	0.00130				1	730.99	0	
730	Arachidic Acid (20:0) (% (w/w))	0619	0.00000	0.00000				1	730.01	4	
732	Gondoic Acid (11c-20:1) (% (w/w))	0226	0.00885	0.00030				2	732.99	0	
732	Gondoic Acid (11c-20:1) (% (w/w))	0619	0.02650	0.00100				2	732.01	0	
736	Arachidonic Acid (5c,8c,11c,14c-20:4) (% (w/v	0676	0.00300	0.00000				1	736.99	0	
736	Arachidonic Acid (5c,8c,11c,14c-20:4) (% (w/v	0619	0.00000	0.00000				1	736.01	4	
738	Mead Acid (11c,14c,17c-20:3) (% (w/w))	0619	0.00000	0.00000				0	738.01	4	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c	0619	0.00000	0.00000				0	740.01	4	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c	0226	0.00000	0.00000				0	740.99	4	
742	Behenic Acid (22:0) (% (w/w))	0226	0.00740	0.00060				1	742.99	0	
742	Behenic Acid (22:0) (% (w/w))	0164	< 0.02					1	742.99	5	
744	Erucic Acid (13c-22:1) (% (w/w))	0619	0.00000	0.00000				0	744.01	4	
744	Erucic Acid (13c-22:1) (% (w/w))	0226	0.00000	0.00000				0	744.99	4	
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,1	0619	0.00000	0.00000				0	746.01	4	
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,1	0226	0.00000	0.00000				0	746.99	4	
748	Lignoceric Acid (24:0) (% (w/w))	0226	0.01225	0.00250				1	748.99	0	
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16	0619	0.00000	0.00000				0	750.01	4	
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16	0226	0.00000	0.00000				0	750.99	4	
752	Nervonic Acid (24:1) isomers (% (w/w))	0619	0.00000	0.00000				0	752.01	4	
752	Nervonic Acid (24:1) isomers (% (w/w))	0226	0.00000	0.00000				0	752.99	4	
754	Total n-3 Polyunsaturated (Omega-3) Fatty Ac	0164	0.09000	0.00000				3	754.99	0	
754	Total n-3 Polyunsaturated (Omega-3) Fatty Ac	0676	0.11000	0.00000				3	754.99	0	
754	Total n-3 Polyunsaturated (Omega-3) Fatty Ac	0171	3.2500	0.10000				3	754.99	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Ac	0164	0.87000	0.04000				3	756.99	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Ac	0676	1.0700	0.00000				3	756.99	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Ac	0171	36.900	1.8000				3	756.99	0	
758	Total Saturated Fatty Acids (% (w/w))	0164	0.78000	0.02000				2	758.99	0	
758	Total Saturated Fatty Acids (% (w/w))	0171	27.000	1.8000				2	758.99	0	
762	Total Monounsaturated Fatty Acids (% (w/w))	0164	0.71000	0.02000				2	762.99	0	
762	Total Monounsaturated Fatty Acids (% (w/w))	0171	29.050	0.30000				2	762.99	0	
766	Total Polyunsaturated Fatty Acids (% (w/w))	0164	0.97000	0.04000				2	766.99	0	
766	Total Polyunsaturated Fatty Acids (% (w/w))	0171	40.150	1.9000				2	766.99	0	
772	Total Fatty Acids (% (w/w))	0226	2.3798	0.07300				2	772.99	0	
772	Total Fatty Acids (% (w/w))	0164	2.4900	0.04000				2	772.99	0	

Notes: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2 and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = Rejected for duplicates too far apart, 2 = Rejected as outlier, 5 = A reporting limit and 4 = zeros submitted as

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			

values. Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). Flag 3 indicates not used in statistics.