

Animal Feed Scheme
Poultry Feed

Labs Reporting: 195
Analytes Reported 112
Issue Date : 07/31/2019

Test Material Code # 201926

Analyte Proficiency Testing Report

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
000	Urea (%)	0278	0.1000	0.0000				1		000.02	0
001	Loss on Drying (%)	0618	7.300	0.0200	8.822	0.3382	0.1141	84	-4.50	001.99	0
001	Loss on Drying (%)	0505	7.920	0.1000	8.822	0.3382	0.1141	84	-2.67	001.07	0
001	Loss on Drying (%)	0345	8.000	0.0800	8.822	0.3382	0.1141	84	-2.43	001.07	0
001	Loss on Drying (%)	0226	8.105	0.3700	8.822	0.3382	0.1141	84	-2.12	001.07	0
001	Loss on Drying (%)	0038	8.110	0.1000	8.822	0.3382	0.1141	84	-2.11	001.07	0
001	Loss on Drying (%)	2259	8.176	0.1180	8.822	0.3382	0.1141	84	-1.91	001.07	0
001	Loss on Drying (%)	0683	8.220	0.2000	8.822	0.3382	0.1141	84	-1.78	001.07	0
001	Loss on Drying (%)	0675	8.270	0.0600	8.822	0.3382	0.1141	84	-1.63	001.07	0
001	Loss on Drying (%)	0028	8.320	0.1800	8.822	0.3382	0.1141	84	-1.48	001.07	0
001	Loss on Drying (%)	2155	8.340	0.3000	8.822	0.3382	0.1141	84	-1.43	001.99	0
001	Loss on Drying (%)	0019	8.350	0.1000	8.822	0.3382	0.1141	84	-1.40	001.07	0
001	Loss on Drying (%)	0948	8.375	0.0700	8.822	0.3382	0.1141	84	-1.32	001.99	0
001	Loss on Drying (%)	0510	8.450	0.1000	8.822	0.3382	0.1141	84	-1.10	001.99	0
001	Loss on Drying (%)	0676	8.475	0.1300	8.822	0.3382	0.1141	84	-1.03	001.99	0
001	Loss on Drying (%)	0849	8.490	0.0600	8.822	0.3382	0.1141	84	-0.98	001.07	0
001	Loss on Drying (%)	0353	8.505	0.2500	8.822	0.3382	0.1141	84	-0.94	001.07	0
001	Loss on Drying (%)	2190	8.515	0.1900	8.822	0.3382	0.1141	84	-0.91	001.07	0
001	Loss on Drying (%)	0142	8.525	0.0300	8.822	0.3382	0.1141	84	-0.88	001.07	0
001	Loss on Drying (%)	2018	8.535	0.0510	8.822	0.3382	0.1141	84	-0.85	001.99	0
001	Loss on Drying (%)	0889	8.545	0.0900	8.822	0.3382	0.1141	84	-0.82	001.99	0
001	Loss on Drying (%)	2246	8.560	0.1600	8.822	0.3382	0.1141	84	-0.77	001.99	0
001	Loss on Drying (%)	0083	8.565	0.0900	8.822	0.3382	0.1141	84	-0.76	001.07	0
001	Loss on Drying (%)	0034	8.565	0.2100	8.822	0.3382	0.1141	84	-0.76	001.07	0
001	Loss on Drying (%)	0187	8.565	0.0300	8.822	0.3382	0.1141	84	-0.76	001.07	0
001	Loss on Drying (%)	0590	8.645	0.3100	8.822	0.3382	0.1141	84	-0.52	001.07	0
001	Loss on Drying (%)	0504	8.655	0.1900	8.822	0.3382	0.1141	84	-0.49	001.00	0
001	Loss on Drying (%)	2126	8.685	0.0300	8.822	0.3382	0.1141	84	-0.41	001.99	0
001	Loss on Drying (%)	0853	8.700	0.0000	8.822	0.3382	0.1141	84	-0.36	001.07	0
001	Loss on Drying (%)	2009	8.730	0.0065	8.822	0.3382	0.1141	84	-0.27	001.07	0
001	Loss on Drying (%)	0720	8.760	0.0800	8.822	0.3382	0.1141	84	-0.18	001.99	0
001	Loss on Drying (%)	0171	8.775	0.0100	8.822	0.3382	0.1141	84	-0.14	001.07	0
001	Loss on Drying (%)	0918	8.775	0.1700	8.822	0.3382	0.1141	84	-0.14	001.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
001	Loss on Drying (%)	0868	8.785	0.0100	8.822	0.3382	0.1141	84	-0.11	001.03	0
001	Loss on Drying (%)	0089	8.790	0.0200	8.822	0.3382	0.1141	84	-0.09	001.07	0
001	Loss on Drying (%)	2242	8.800	0.4000	8.822	0.3382	0.1141	84	-0.06	001.99	0
001	Loss on Drying (%)	0612	8.810	0.0000	8.822	0.3382	0.1141	84	-0.04	001.99	0
001	Loss on Drying (%)	0169	8.815	0.0300	8.822	0.3382	0.1141	84	-0.02	001.00	0
001	Loss on Drying (%)	0852	8.825	0.0100	8.822	0.3382	0.1141	84	0.01	001.99	0
001	Loss on Drying (%)	0610	8.830	0.0200	8.822	0.3382	0.1141	84	0.02	001.05	0
001	Loss on Drying (%)	2109	8.835	0.3100	8.822	0.3382	0.1141	84	0.04	001.07	0
001	Loss on Drying (%)	0910	8.850	0.1000	8.822	0.3382	0.1141	84	0.08	001.99	0
001	Loss on Drying (%)	0066	8.860	0.1200	8.822	0.3382	0.1141	84	0.11	001.07	0
001	Loss on Drying (%)	0682	8.860	0.0000	8.822	0.3382	0.1141	84	0.11	001.07	0
001	Loss on Drying (%)	0035	8.865	0.0300	8.822	0.3382	0.1141	84	0.13	001.07	0
001	Loss on Drying (%)	0560	8.875	0.2300	8.822	0.3382	0.1141	84	0.16	001.99	0
001	Loss on Drying (%)	2062	8.879	0.0057	8.822	0.3382	0.1141	84	0.17	001.03	0
001	Loss on Drying (%)	0297	8.885	0.0500	8.822	0.3382	0.1141	84	0.19	001.07	0
001	Loss on Drying (%)	2192	8.950	0.3200	8.822	0.3382	0.1141	84	0.38	001.07	0
001	Loss on Drying (%)	0074	8.960	0.0000	8.822	0.3382	0.1141	84	0.41	001.07	0
001	Loss on Drying (%)	0037	8.975	0.0700	8.822	0.3382	0.1141	84	0.45	001.99	0
001	Loss on Drying (%)	0033	8.975	0.0100	8.822	0.3382	0.1141	84	0.45	001.07	0
001	Loss on Drying (%)	0662	8.975	0.1686	8.822	0.3382	0.1141	84	0.45	001.99	0
001	Loss on Drying (%)	0098	8.980	0.1400	8.822	0.3382	0.1141	84	0.47	001.07	0
001	Loss on Drying (%)	0309	9.000	0.0200	8.822	0.3382	0.1141	84	0.53	001.00	0
001	Loss on Drying (%)	0689	9.000	0.0000	8.822	0.3382	0.1141	84	0.53	001.07	0
001	Loss on Drying (%)	2053	9.000	0.1600	8.822	0.3382	0.1141	84	0.53	001.07	0
001	Loss on Drying (%)	2059	9.000	0.0200	8.822	0.3382	0.1141	84	0.53	001.07	0
001	Loss on Drying (%)	0695	9.001	0.0990	8.822	0.3382	0.1141	84	0.53	001.07	0
001	Loss on Drying (%)	0049	9.020	0.3000	8.822	0.3382	0.1141	84	0.59	001.07	0
001	Loss on Drying (%)	0278	9.025	0.0900	8.822	0.3382	0.1141	84	0.60	001.07	0
001	Loss on Drying (%)	2264	9.034	0.1074	8.822	0.3382	0.1141	84	0.63	001.07	0
001	Loss on Drying (%)	0656	9.055	0.0900	8.822	0.3382	0.1141	84	0.69	001.99	0
001	Loss on Drying (%)	0629	9.075	0.0300	8.822	0.3382	0.1141	84	0.75	001.99	0
001	Loss on Drying (%)	2268	9.075	0.2500	8.822	0.3382	0.1141	84	0.75	001.99	0
001	Loss on Drying (%)	0027	9.081	0.0090	8.822	0.3382	0.1141	84	0.76	001.00	0
001	Loss on Drying (%)	0912	9.085	0.0700	8.822	0.3382	0.1141	84	0.78	001.99	0
001	Loss on Drying (%)	0619	9.090	0.0200	8.822	0.3382	0.1141	84	0.79	001.03	0
001	Loss on Drying (%)	0872	9.095	0.1100	8.822	0.3382	0.1141	84	0.81	001.07	0
001	Loss on Drying (%)	0581	9.105	0.1100	8.822	0.3382	0.1141	84	0.84	001.07	0
001	Loss on Drying (%)	0407	9.110	0.0983	8.822	0.3382	0.1141	84	0.85	001.07	0
001	Loss on Drying (%)	0878	9.115	0.1100	8.822	0.3382	0.1141	84	0.87	001.03	0
001	Loss on Drying (%)	0571	9.125	0.0300	8.822	0.3382	0.1141	84	0.90	001.07	0
001	Loss on Drying (%)	0843	9.130	0.1600	8.822	0.3382	0.1141	84	0.91	001.03	0
001	Loss on Drying (%)	0843	9.130	0.1600	8.822	0.3382	0.1141	84	0.91	001.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
001	Loss on Drying (%)	0876	9.150	0.0200	8.822	0.3382	0.1141	84	0.97	001.07	0
001	Loss on Drying (%)	0357	9.156	0.2226	8.822	0.3382	0.1141	84	0.99	001.99	0
001	Loss on Drying (%)	0148	9.160	0.0200	8.822	0.3382	0.1141	84	1.00	001.07	0
001	Loss on Drying (%)	0019	9.160	0.0400	8.822	0.3382	0.1141	84	1.00	001.99	0
001	Loss on Drying (%)	0644	9.210	0.0200	8.822	0.3382	0.1141	84	1.15	001.05	0
001	Loss on Drying (%)	0045	9.270	0.1000	8.822	0.3382	0.1141	84	1.32	001.07	0
001	Loss on Drying (%)	0893	9.300	0.0200	8.822	0.3382	0.1141	84	1.41	001.03	0
001	Loss on Drying (%)	2081	9.390	0.6600	8.822	0.3382	0.1141	84	1.68	001.99	0
001	Loss on Drying (%)	0693	9.612	0.4300	8.822	0.3382	0.1141	84	2.34	001.07	0
001	Loss on Drying (%)	0529	9.740	0.0800	8.822	0.3382	0.1141	84	2.71	001.07	0
001	Loss on Drying (%)	0003	9.960	2.040	8.822	0.3382	0.1141	84	3.37	001.07	1
001	Loss on Drying (%)	0366	10.10	1.000	8.822	0.3382	0.1141	84	3.78	001.07	1
002	Protein, Crude (%)	0889	20.51	0.1200	21.91	0.3331	0.1835	189	-4.21	002.11	0
002	Protein, Crude (%)	2128	20.53	0.2020	21.91	0.3331	0.1835	189	-4.15	002.03	0
002	Protein, Crude (%)	0852	20.54	0.0900	21.91	0.3331	0.1835	189	-4.13	002.05	0
002	Protein, Crude (%)	0852	20.99	0.0200	21.91	0.3331	0.1835	189	-2.77	002.11	0
002	Protein, Crude (%)	2248	21.01	0.0100	21.91	0.3331	0.1835	189	-2.72	002.05	0
002	Protein, Crude (%)	0345	21.11	0.1800	21.91	0.3331	0.1835	189	-2.41	002.06	0
002	Protein, Crude (%)	0119	21.21	0.5670	21.91	0.3331	0.1835	189	-2.09	002.06	0
002	Protein, Crude (%)	0169	21.26	0.1100	21.91	0.3331	0.1835	189	-1.97	002.02	0
002	Protein, Crude (%)	0939	21.28	0.2800	21.91	0.3331	0.1835	189	-1.90	002.05	0
002	Protein, Crude (%)	0720	21.29	0.0200	21.91	0.3331	0.1835	189	-1.87	002.06	0
002	Protein, Crude (%)	2259	21.30	0.6078	21.91	0.3331	0.1835	189	-1.84	002.01	0
002	Protein, Crude (%)	0581	21.33	0.0600	21.91	0.3331	0.1835	189	-1.75	002.06	0
002	Protein, Crude (%)	0536	21.34	0.0400	21.91	0.3331	0.1835	189	-1.72	002.05	0
002	Protein, Crude (%)	0164	21.35	0.1900	21.91	0.3331	0.1835	189	-1.70	002.01	0
002	Protein, Crude (%)	2155	21.38	0.2830	21.91	0.3331	0.1835	189	-1.58	002.06	0
002	Protein, Crude (%)	0955	21.40	0.2000	21.91	0.3331	0.1835	189	-1.54	002.06	0
002	Protein, Crude (%)	2264	21.46	0.0010	21.91	0.3331	0.1835	189	-1.34	002.05	0
002	Protein, Crude (%)	2190	21.47	0.1100	21.91	0.3331	0.1835	189	-1.34	002.05	0
002	Protein, Crude (%)	0874	21.47	0.4400	21.91	0.3331	0.1835	189	-1.33	002.04	0
002	Protein, Crude (%)	0598	21.48	0.1200	21.91	0.3331	0.1835	189	-1.30	002.06	0
002	Protein, Crude (%)	0893	21.49	0.0300	21.91	0.3331	0.1835	189	-1.28	002.05	0
002	Protein, Crude (%)	0036	21.49	0.0858	21.91	0.3331	0.1835	189	-1.27	002.02	0
002	Protein, Crude (%)	0629	21.49	0.0800	21.91	0.3331	0.1835	189	-1.27	002.10	0
002	Protein, Crude (%)	0955	21.50	0.4000	21.91	0.3331	0.1835	189	-1.24	002.05	0
002	Protein, Crude (%)	0650	21.50	0.2000	21.91	0.3331	0.1835	189	-1.24	002.06	0
002	Protein, Crude (%)	2193	21.54	0.1700	21.91	0.3331	0.1835	189	-1.13	002.01	0
002	Protein, Crude (%)	0904	21.54	0.1900	21.91	0.3331	0.1835	189	-1.13	002.06	0
002	Protein, Crude (%)	0510	21.55	0.1000	21.91	0.3331	0.1835	189	-1.09	002.06	0
002	Protein, Crude (%)	2251	21.56	0.0300	21.91	0.3331	0.1835	189	-1.07	002.06	0
002	Protein, Crude (%)	0939	21.57	0.1200	21.91	0.3331	0.1835	189	-1.03	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein, Crude (%)	2109	21.58	0.2800	21.91	0.3331	0.1835	189	-1.00	002.05	0
002	Protein, Crude (%)	0354	21.59	0.0400	21.91	0.3331	0.1835	189	-0.97	002.05	0
002	Protein, Crude (%)	0662	21.59	0.1875	21.91	0.3331	0.1835	189	-0.96	002.01	0
002	Protein, Crude (%)	0910	21.60	0.4000	21.91	0.3331	0.1835	189	-0.94	002.01	0
002	Protein, Crude (%)	0675	21.63	0.0500	21.91	0.3331	0.1835	189	-0.86	002.05	0
002	Protein, Crude (%)	0539	21.63	0.7900	21.91	0.3331	0.1835	189	-0.86	002.06	0
002	Protein, Crude (%)	0066	21.65	0.1180	21.91	0.3331	0.1835	189	-0.80	002.02	0
002	Protein, Crude (%)	0652	21.65	0.1000	21.91	0.3331	0.1835	189	-0.79	002.01	0
002	Protein, Crude (%)	0689	21.65	0.1000	21.91	0.3331	0.1835	189	-0.79	002.05	0
002	Protein, Crude (%)	2073	21.65	0.0200	21.91	0.3331	0.1835	189	-0.79	002.05	0
002	Protein, Crude (%)	2199	21.65	0.1000	21.91	0.3331	0.1835	189	-0.79	002.05	0
002	Protein, Crude (%)	0028	21.65	0.4600	21.91	0.3331	0.1835	189	-0.79	002.06	0
002	Protein, Crude (%)	0529	21.65	0.8200	21.91	0.3331	0.1835	189	-0.79	002.06	0
002	Protein, Crude (%)	0695	21.68	0.0720	21.91	0.3331	0.1835	189	-0.69	002.06	0
002	Protein, Crude (%)	2006	21.70	0.1400	21.91	0.3331	0.1835	189	-0.64	002.05	0
002	Protein, Crude (%)	0014	21.70	0.2000	21.91	0.3331	0.1835	189	-0.64	002.06	0
002	Protein, Crude (%)	2150	21.71	0.0200	21.91	0.3331	0.1835	189	-0.61	002.06	0
002	Protein, Crude (%)	0011	21.72	0.3000	21.91	0.3331	0.1835	189	-0.58	002.06	0
002	Protein, Crude (%)	0968	21.73	0.0540	21.91	0.3331	0.1835	189	-0.53	002.01	0
002	Protein, Crude (%)	0505	21.74	0.0500	21.91	0.3331	0.1835	189	-0.53	002.06	0
002	Protein, Crude (%)	2022	21.74	0.1000	21.91	0.3331	0.1835	189	-0.52	002.05	0
002	Protein, Crude (%)	0047	21.74	0.6000	21.91	0.3331	0.1835	189	-0.52	002.06	0
002	Protein, Crude (%)	0619	21.75	0.1000	21.91	0.3331	0.1835	189	-0.49	002.05	0
002	Protein, Crude (%)	0881	21.75	0.0180	21.91	0.3331	0.1835	189	-0.49	002.05	0
002	Protein, Crude (%)	0265	21.75	0.1000	21.91	0.3331	0.1835	189	-0.49	002.06	0
002	Protein, Crude (%)	0520	21.75	0.1000	21.91	0.3331	0.1835	189	-0.49	002.06	0
002	Protein, Crude (%)	0870	21.75	0.0100	21.91	0.3331	0.1835	189	-0.48	002.01	0
002	Protein, Crude (%)	0656	21.76	0.1700	21.91	0.3331	0.1835	189	-0.47	002.01	0
002	Protein, Crude (%)	2129	21.76	0.1250	21.91	0.3331	0.1835	189	-0.47	002.06	0
002	Protein, Crude (%)	0003	21.76	0.1000	21.91	0.3331	0.1835	189	-0.46	002.06	0
002	Protein, Crude (%)	0144	21.76	0.0600	21.91	0.3331	0.1835	189	-0.46	002.06	0
002	Protein, Crude (%)	0692	21.76	0.5200	21.91	0.3331	0.1835	189	-0.46	002.06	0
002	Protein, Crude (%)	0171	21.78	0.0200	21.91	0.3331	0.1835	189	-0.40	002.06	0
002	Protein, Crude (%)	0885	21.80	0.0283	21.91	0.3331	0.1835	189	-0.35	002.05	0
002	Protein, Crude (%)	2192	21.81	0.1300	21.91	0.3331	0.1835	189	-0.32	002.05	0
002	Protein, Crude (%)	2188	21.81	0.0000	21.91	0.3331	0.1835	189	-0.31	002.01	0
002	Protein, Crude (%)	2181	21.81	0.0200	21.91	0.3331	0.1835	189	-0.31	002.05	0
002	Protein, Crude (%)	0541	21.81	0.4000	21.91	0.3331	0.1835	189	-0.31	002.06	0
002	Protein, Crude (%)	0849	21.81	0.0170	21.91	0.3331	0.1835	189	-0.30	002.05	0
002	Protein, Crude (%)	2062	21.81	0.2509	21.91	0.3331	0.1835	189	-0.30	002.05	0
002	Protein, Crude (%)	0074	21.82	0.1200	21.91	0.3331	0.1835	189	-0.28	002.06	0
002	Protein, Crude (%)	0589	21.82	0.2400	21.91	0.3331	0.1835	189	-0.28	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein, Crude (%)	0626	21.83	0.1560	21.91	0.3331	0.1835	189	-0.25	002.06	0
002	Protein, Crude (%)	2196	21.84	0.0000	21.91	0.3331	0.1835	189	-0.22	002.01	0
002	Protein, Crude (%)	0098	21.85	0.1000	21.91	0.3331	0.1835	189	-0.19	002.06	0
002	Protein, Crude (%)	0610	21.85	0.1000	21.91	0.3331	0.1835	189	-0.19	002.06	0
002	Protein, Crude (%)	0880	21.85	0.5000	21.91	0.3331	0.1835	189	-0.19	002.06	0
002	Protein, Crude (%)	0683	21.86	0.0700	21.91	0.3331	0.1835	189	-0.17	002.05	0
002	Protein, Crude (%)	0953	21.86	0.0300	21.91	0.3331	0.1835	189	-0.17	002.06	0
002	Protein, Crude (%)	0037	21.87	0.4500	21.91	0.3331	0.1835	189	-0.14	002.06	0
002	Protein, Crude (%)	0354	21.87	0.0200	21.91	0.3331	0.1835	189	-0.13	002.06	0
002	Protein, Crude (%)	0676	21.88	0.2500	21.91	0.3331	0.1835	189	-0.11	002.06	0
002	Protein, Crude (%)	0870	21.88	0.2638	21.91	0.3331	0.1835	189	-0.09	002.06	0
002	Protein, Crude (%)	0868	21.89	0.0100	21.91	0.3331	0.1835	189	-0.08	002.05	0
002	Protein, Crude (%)	0202	21.89	0.3500	21.91	0.3331	0.1835	189	-0.08	002.06	0
002	Protein, Crude (%)	0941	21.89	0.0200	21.91	0.3331	0.1835	189	-0.07	002.05	0
002	Protein, Crude (%)	0968	21.89	0.1320	21.91	0.3331	0.1835	189	-0.06	002.06	0
002	Protein, Crude (%)	0868	21.90	0.0100	21.91	0.3331	0.1835	189	-0.05	002.06	0
002	Protein, Crude (%)	0610	21.90	0.0000	21.91	0.3331	0.1835	189	-0.04	002.01	0
002	Protein, Crude (%)	0108	21.90	0.0400	21.91	0.3331	0.1835	189	-0.04	002.06	0
002	Protein, Crude (%)	0175	21.90	0.0000	21.91	0.3331	0.1835	189	-0.04	002.06	0
002	Protein, Crude (%)	0294	21.90	0.0600	21.91	0.3331	0.1835	189	-0.04	002.06	0
002	Protein, Crude (%)	0884	21.90	0.2000	21.91	0.3331	0.1835	189	-0.04	002.06	0
002	Protein, Crude (%)	0089	21.91	0.0100	21.91	0.3331	0.1835	189	-0.02	002.06	0
002	Protein, Crude (%)	0123	21.91	0.2200	21.91	0.3331	0.1835	189	-0.01	002.06	0
002	Protein, Crude (%)	0309	21.91	0.2825	21.91	0.3331	0.1835	189	0.01	002.06	0
002	Protein, Crude (%)	0878	21.92	0.1400	21.91	0.3331	0.1835	189	0.02	002.05	0
002	Protein, Crude (%)	0948	21.93	0.0900	21.91	0.3331	0.1835	189	0.04	002.06	0
002	Protein, Crude (%)	0687	21.93	0.1500	21.91	0.3331	0.1835	189	0.04	002.06	0
002	Protein, Crude (%)	0964	21.93	0.0090	21.91	0.3331	0.1835	189	0.05	002.06	0
002	Protein, Crude (%)	2198	21.93	0.1000	21.91	0.3331	0.1835	189	0.05	002.01	0
002	Protein, Crude (%)	0948	21.93	0.0600	21.91	0.3331	0.1835	189	0.05	002.05	0
002	Protein, Crude (%)	0353	21.94	0.1200	21.91	0.3331	0.1835	189	0.08	002.06	0
002	Protein, Crude (%)	0685	21.94	0.1400	21.91	0.3331	0.1835	189	0.08	002.01	0
002	Protein, Crude (%)	2009	21.95	0.0053	21.91	0.3331	0.1835	189	0.10	002.05	0
002	Protein, Crude (%)	0872	21.95	0.1755	21.91	0.3331	0.1835	189	0.11	002.06	0
002	Protein, Crude (%)	0619	21.95	0.1000	21.91	0.3331	0.1835	189	0.11	002.06	0
002	Protein, Crude (%)	0682	21.95	0.0000	21.91	0.3331	0.1835	189	0.11	002.06	0
002	Protein, Crude (%)	2033	21.95	0.1800	21.91	0.3331	0.1835	189	0.11	002.06	0
002	Protein, Crude (%)	0032	21.95	0.3000	21.91	0.3331	0.1835	189	0.11	002.06	0
002	Protein, Crude (%)	0848	21.96	0.0400	21.91	0.3331	0.1835	189	0.14	002.06	0
002	Protein, Crude (%)	0100	21.97	0.0900	21.91	0.3331	0.1835	189	0.16	002.06	0
002	Protein, Crude (%)	0425	21.98	0.0300	21.91	0.3331	0.1835	189	0.19	002.06	0
002	Protein, Crude (%)	0675	22.00	0.0200	21.91	0.3331	0.1835	189	0.26	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
002	Protein, Crude (%)	0910	22.00	0.2000	21.91	0.3331	0.1835	189	0.26	002.06	0
002	Protein, Crude (%)	0956	22.00	0.0000	21.91	0.3331	0.1835	189	0.26	002.06	0
002	Protein, Crude (%)	0164	22.01	0.2500	21.91	0.3331	0.1835	189	0.28	002.06	0
002	Protein, Crude (%)	0857	22.01	0.2100	21.91	0.3331	0.1835	189	0.28	002.06	0
002	Protein, Crude (%)	0297	22.01	0.1400	21.91	0.3331	0.1835	189	0.29	002.06	0
002	Protein, Crude (%)	0142	22.02	0.1100	21.91	0.3331	0.1835	189	0.31	002.06	0
002	Protein, Crude (%)	2053	22.02	0.5300	21.91	0.3331	0.1835	189	0.31	002.06	0
002	Protein, Crude (%)	0098	22.02	0.2900	21.91	0.3331	0.1835	189	0.31	002.08	0
002	Protein, Crude (%)	0036	22.03	0.0400	21.91	0.3331	0.1835	189	0.35	002.06	0
002	Protein, Crude (%)	0574	22.03	0.0600	21.91	0.3331	0.1835	189	0.35	002.06	0
002	Protein, Crude (%)	0049	22.05	0.0500	21.91	0.3331	0.1835	189	0.40	002.06	0
002	Protein, Crude (%)	0190	22.05	0.4300	21.91	0.3331	0.1835	189	0.40	002.06	0
002	Protein, Crude (%)	0010	22.05	0.1000	21.91	0.3331	0.1835	189	0.41	002.06	0
002	Protein, Crude (%)	0017	22.05	0.1000	21.91	0.3331	0.1835	189	0.41	002.06	0
002	Protein, Crude (%)	0226	22.05	0.1000	21.91	0.3331	0.1835	189	0.41	002.06	0
002	Protein, Crude (%)	0278	22.05	0.1000	21.91	0.3331	0.1835	189	0.41	002.06	0
002	Protein, Crude (%)	0366	22.05	0.5000	21.91	0.3331	0.1835	189	0.41	002.06	0
002	Protein, Crude (%)	0508	22.05	0.2820	21.91	0.3331	0.1835	189	0.42	002.06	0
002	Protein, Crude (%)	0038	22.06	0.3960	21.91	0.3331	0.1835	189	0.44	002.06	0
002	Protein, Crude (%)	0357	22.06	0.1250	21.91	0.3331	0.1835	189	0.45	002.06	0
002	Protein, Crude (%)	0035	22.07	0.0200	21.91	0.3331	0.1835	189	0.47	002.06	0
002	Protein, Crude (%)	2076	22.07	0.3460	21.91	0.3331	0.1835	189	0.48	002.06	0
002	Protein, Crude (%)	0571	22.08	0.0500	21.91	0.3331	0.1835	189	0.49	002.06	0
002	Protein, Crude (%)	0592	22.08	0.3200	21.91	0.3331	0.1835	189	0.50	002.06	0
002	Protein, Crude (%)	0045	22.10	0.4000	21.91	0.3331	0.1835	189	0.56	002.06	0
002	Protein, Crude (%)	0083	22.11	0.1900	21.91	0.3331	0.1835	189	0.58	002.06	0
002	Protein, Crude (%)	2184	22.11	0.2600	21.91	0.3331	0.1835	189	0.59	002.06	0
002	Protein, Crude (%)	0644	22.12	0.0900	21.91	0.3331	0.1835	189	0.61	002.06	0
002	Protein, Crude (%)	2168	22.12	0.0900	21.91	0.3331	0.1835	189	0.61	002.06	0
002	Protein, Crude (%)	0897	22.12	0.0000	21.91	0.3331	0.1835	189	0.62	002.05	0
002	Protein, Crude (%)	0263	22.14	0.1500	21.91	0.3331	0.1835	189	0.67	002.06	0
002	Protein, Crude (%)	0504	22.14	0.1200	21.91	0.3331	0.1835	189	0.68	002.04	0
002	Protein, Crude (%)	2059	22.14	0.0688	21.91	0.3331	0.1835	189	0.69	002.06	0
002	Protein, Crude (%)	0004	22.15	0.3800	21.91	0.3331	0.1835	189	0.71	002.06	0
002	Protein, Crude (%)	0026	22.16	0.4060	21.91	0.3331	0.1835	189	0.74	002.06	0
002	Protein, Crude (%)	0242	22.16	0.1800	21.91	0.3331	0.1835	189	0.74	002.06	0
002	Protein, Crude (%)	2126	22.17	0.0700	21.91	0.3331	0.1835	189	0.76	002.06	0
002	Protein, Crude (%)	0553	22.17	0.3700	21.91	0.3331	0.1835	189	0.76	002.11	0
002	Protein, Crude (%)	0001	22.17	0.0800	21.91	0.3331	0.1835	189	0.77	002.06	0
002	Protein, Crude (%)	0590	22.18	0.0600	21.91	0.3331	0.1835	189	0.80	002.06	0
002	Protein, Crude (%)	0298	22.20	0.0100	21.91	0.3331	0.1835	189	0.85	002.06	0
002	Protein, Crude (%)	2089	22.20	0.1000	21.91	0.3331	0.1835	189	0.86	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0843	22.23	0.4500	21.91	0.3331	0.1835	189	0.94	002.06	0
002	Protein, Crude (%)	0229	22.23	0.1200	21.91	0.3331	0.1835	189	0.95	002.06	0
002	Protein, Crude (%)	0511	22.24	0.3100	21.91	0.3331	0.1835	189	0.97	002.06	0
002	Protein, Crude (%)	0407	22.25	0.0990	21.91	0.3331	0.1835	189	1.02	002.06	0
002	Protein, Crude (%)	2171	22.25	0.2820	21.91	0.3331	0.1835	189	1.02	002.06	0
002	Protein, Crude (%)	0859	22.26	0.2430	21.91	0.3331	0.1835	189	1.03	002.06	0
002	Protein, Crude (%)	0656	22.26	0.3800	21.91	0.3331	0.1835	189	1.04	002.06	0
002	Protein, Crude (%)	2081	22.28	0.8700	21.91	0.3331	0.1835	189	1.09	002.06	0
002	Protein, Crude (%)	0148	22.28	0.0800	21.91	0.3331	0.1835	189	1.10	002.06	0
002	Protein, Crude (%)	0504	22.30	0.1000	21.91	0.3331	0.1835	189	1.16	002.06	0
002	Protein, Crude (%)	0034	22.31	0.5570	21.91	0.3331	0.1835	189	1.21	002.06	0
002	Protein, Crude (%)	0693	22.33	0.3350	21.91	0.3331	0.1835	189	1.25	002.06	0
002	Protein, Crude (%)	0358	22.37	1.000	21.91	0.3331	0.1835	189	1.38	002.06	0
002	Protein, Crude (%)	0695	22.37	0.1000	21.91	0.3331	0.1835	189	1.38	002.11	0
002	Protein, Crude (%)	2242	22.40	0.0100	21.91	0.3331	0.1835	189	1.45	002.06	0
002	Protein, Crude (%)	0187	22.41	0.1000	21.91	0.3331	0.1835	189	1.50	002.04	0
002	Protein, Crude (%)	0019	22.46	0.4100	21.91	0.3331	0.1835	189	1.63	002.06	0
002	Protein, Crude (%)	0918	22.49	0.5100	21.91	0.3331	0.1835	189	1.72	002.06	0
002	Protein, Crude (%)	0876	22.52	0.0200	21.91	0.3331	0.1835	189	1.83	002.06	0
002	Protein, Crude (%)	0720	22.54	0.1200	21.91	0.3331	0.1835	189	1.89	002.11	0
002	Protein, Crude (%)	2213	22.55	0.1300	21.91	0.3331	0.1835	189	1.90	002.06	0
002	Protein, Crude (%)	2246	22.58	0.0100	21.91	0.3331	0.1835	189	1.99	002.06	0
002	Protein, Crude (%)	0618	22.58	0.0670	21.91	0.3331	0.1835	189	1.99	002.06	0
002	Protein, Crude (%)	0726	22.63	0.3000	21.91	0.3331	0.1835	189	2.16	002.06	0
002	Protein, Crude (%)	2018	22.91	0.1500	21.91	0.3331	0.1835	189	2.98	002.06	0
002	Protein, Crude (%)	2109	22.92	0.0100	21.91	0.3331	0.1835	189	3.01	002.06	0
002	Protein, Crude (%)	2128	23.16	0.2380	21.91	0.3331	0.1835	189	3.76	002.11	0
002	Protein, Crude (%)	2268	23.39	1.550	21.91	0.3331	0.1835	189	4.42	002.06	0
002	Protein, Crude (%)	0553	23.91	0.2600	21.91	0.3331	0.1835	189	6.00	002.06	0
002	Protein, Crude (%)	0297	24.40	0.0300	21.91	0.3331	0.1835	189	7.46	002.11	0
002	Protein, Crude (%)	0011	24.70	0.0000	21.91	0.3331	0.1835	189	8.37	002.11	0
002	Protein, Crude (%)	0612	25.73	0.0900	21.91	0.3331	0.1835	189	11.45	002.06	0
002	Protein, Crude (%)	0853	26.09	0.3700	21.91	0.3331	0.1835	189	12.53	002.06	0
002	Protein, Crude (%)	0062	25.15	4.550	21.91	0.3331	0.1835	189	9.71	002.06	1
002	Protein, Crude (%)	0889	16.42	10.91	21.91	0.3331	0.1835	189	-16.50	002.01	2
003	Fat, Crude (%)	0843	2.650	0.3000	3.709	0.2218	0.1126	151	-4.78	003.14	0
003	Fat, Crude (%)	0574	2.685	0.2900	3.709	0.2218	0.1126	151	-4.62	003.14	0
003	Fat, Crude (%)	0003	2.950	0.2200	3.709	0.2218	0.1126	151	-3.42	003.14	0
003	Fat, Crude (%)	2246	3.055	0.0300	3.709	0.2218	0.1126	151	-2.95	003.13	0
003	Fat, Crude (%)	0357	3.200	0.4000	3.709	0.2218	0.1126	151	-2.30	003.14	0
003	Fat, Crude (%)	2129	3.220	0.6200	3.709	0.2218	0.1126	151	-2.21	003.14	0
003	Fat, Crude (%)	0175	3.245	0.1500	3.709	0.2218	0.1126	151	-2.09	003.14	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat, Crude (%)	0004	3.265	0.0500	3.709	0.2218	0.1126	151	-2.00	003.09	0
003	Fat, Crude (%)	0504	3.270	0.4800	3.709	0.2218	0.1126	151	-1.98	003.01	0
003	Fat, Crude (%)	0968	3.381	0.1630	3.709	0.2218	0.1126	151	-1.48	003.14	0
003	Fat, Crude (%)	0619	3.385	0.0500	3.709	0.2218	0.1126	151	-1.46	003.10	0
003	Fat, Crude (%)	0045	3.395	0.4700	3.709	0.2218	0.1126	151	-1.42	003.10	0
003	Fat, Crude (%)	0912	3.400	0.1000	3.709	0.2218	0.1126	151	-1.39	003.14	0
003	Fat, Crude (%)	0893	3.405	0.0500	3.709	0.2218	0.1126	151	-1.37	003.12	0
003	Fat, Crude (%)	2199	3.450	0.1000	3.709	0.2218	0.1126	151	-1.17	003.12	0
003	Fat, Crude (%)	0536	3.450	0.1000	3.709	0.2218	0.1126	151	-1.17	003.99	0
003	Fat, Crude (%)	0720	3.475	0.0100	3.709	0.2218	0.1126	151	-1.06	003.11	0
003	Fat, Crude (%)	0683	3.480	0.1400	3.709	0.2218	0.1126	151	-1.03	003.14	0
003	Fat, Crude (%)	2184	3.480	0.0200	3.709	0.2218	0.1126	151	-1.03	003.10	0
003	Fat, Crude (%)	0693	3.483	0.2280	3.709	0.2218	0.1126	151	-1.02	003.10	0
003	Fat, Crude (%)	2073	3.495	0.0300	3.709	0.2218	0.1126	151	-0.97	003.10	0
003	Fat, Crude (%)	0366	3.500	0.2000	3.709	0.2218	0.1126	151	-0.94	003.10	0
003	Fat, Crude (%)	2188	3.500	0.0400	3.709	0.2218	0.1126	151	-0.94	003.10	0
003	Fat, Crude (%)	0298	3.500	0.0200	3.709	0.2218	0.1126	151	-0.94	003.14	0
003	Fat, Crude (%)	0853	3.500	0.4000	3.709	0.2218	0.1126	151	-0.94	003.14	0
003	Fat, Crude (%)	0720	3.505	0.0100	3.709	0.2218	0.1126	151	-0.92	003.10	0
003	Fat, Crude (%)	2018	3.505	0.0300	3.709	0.2218	0.1126	151	-0.92	003.10	0
003	Fat, Crude (%)	0889	3.505	0.0100	3.709	0.2218	0.1126	151	-0.92	003.11	0
003	Fat, Crude (%)	0948	3.505	0.0500	3.709	0.2218	0.1126	151	-0.92	003.14	0
003	Fat, Crude (%)	0876	3.510	0.0400	3.709	0.2218	0.1126	151	-0.90	003.14	0
003	Fat, Crude (%)	0520	3.515	0.0200	3.709	0.2218	0.1126	151	-0.88	003.14	0
003	Fat, Crude (%)	0089	3.515	0.0500	3.709	0.2218	0.1126	151	-0.88	003.14	0
003	Fat, Crude (%)	0108	3.530	0.2600	3.709	0.2218	0.1126	151	-0.81	003.14	0
003	Fat, Crude (%)	0553	3.540	0.0200	3.709	0.2218	0.1126	151	-0.76	003.10	0
003	Fat, Crude (%)	0144	3.545	0.0300	3.709	0.2218	0.1126	151	-0.74	003.14	0
003	Fat, Crude (%)	0868	3.550	0.1200	3.709	0.2218	0.1126	151	-0.72	003.14	0
003	Fat, Crude (%)	0956	3.550	0.1000	3.709	0.2218	0.1126	151	-0.72	003.14	0
003	Fat, Crude (%)	2089	3.555	0.0300	3.709	0.2218	0.1126	151	-0.70	003.13	0
003	Fat, Crude (%)	2022	3.560	0.0600	3.709	0.2218	0.1126	151	-0.67	003.10	0
003	Fat, Crude (%)	0034	3.560	0.4600	3.709	0.2218	0.1126	151	-0.67	003.14	0
003	Fat, Crude (%)	0036	3.563	0.0794	3.709	0.2218	0.1126	151	-0.66	003.10	0
003	Fat, Crude (%)	0033	3.570	0.0800	3.709	0.2218	0.1126	151	-0.63	003.13	0
003	Fat, Crude (%)	2155	3.570	0.0600	3.709	0.2218	0.1126	151	-0.63	003.10	0
003	Fat, Crude (%)	0037	3.575	0.0900	3.709	0.2218	0.1126	151	-0.61	003.14	0
003	Fat, Crude (%)	2062	3.576	0.0593	3.709	0.2218	0.1126	151	-0.60	003.12	0
003	Fat, Crude (%)	0263	3.594	0.0330	3.709	0.2218	0.1126	151	-0.52	003.10	0
003	Fat, Crude (%)	0083	3.595	0.0700	3.709	0.2218	0.1126	151	-0.52	003.14	0
003	Fat, Crude (%)	0511	3.595	0.1900	3.709	0.2218	0.1126	151	-0.52	003.14	0
003	Fat, Crude (%)	0202	3.601	0.3090	3.709	0.2218	0.1126	151	-0.49	003.14	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat, Crude (%)	0345	3.605	0.0900	3.709	0.2218	0.1126	151	-0.47	003.00	0
003	Fat, Crude (%)	0042	3.610	0.2000	3.709	0.2218	0.1126	151	-0.45	003.10	0
003	Fat, Crude (%)	0885	3.623	0.0825	3.709	0.2218	0.1126	151	-0.39	003.00	0
003	Fat, Crude (%)	0881	3.623	0.0280	3.709	0.2218	0.1126	151	-0.39	003.10	0
003	Fat, Crude (%)	0358	3.625	0.1900	3.709	0.2218	0.1126	151	-0.38	003.14	0
003	Fat, Crude (%)	0581	3.630	0.2400	3.709	0.2218	0.1126	151	-0.36	003.14	0
003	Fat, Crude (%)	0598	3.630	0.1400	3.709	0.2218	0.1126	151	-0.36	003.14	0
003	Fat, Crude (%)	0297	3.635	0.0100	3.709	0.2218	0.1126	151	-0.34	003.06	0
003	Fat, Crude (%)	0695	3.644	0.1930	3.709	0.2218	0.1126	151	-0.30	003.10	0
003	Fat, Crude (%)	0948	3.645	0.1500	3.709	0.2218	0.1126	151	-0.29	003.09	0
003	Fat, Crude (%)	0100	3.645	0.1500	3.709	0.2218	0.1126	151	-0.29	003.10	0
003	Fat, Crude (%)	2009	3.647	0.0117	3.709	0.2218	0.1126	151	-0.28	003.10	0
003	Fat, Crude (%)	0062	3.649	0.0160	3.709	0.2218	0.1126	151	-0.27	003.10	0
003	Fat, Crude (%)	0066	3.651	0.0540	3.709	0.2218	0.1126	151	-0.26	003.14	0
003	Fat, Crude (%)	2193	3.660	0.0400	3.709	0.2218	0.1126	151	-0.22	003.10	0
003	Fat, Crude (%)	0123	3.670	0.0800	3.709	0.2218	0.1126	151	-0.18	003.13	0
003	Fat, Crude (%)	0505	3.670	0.1400	3.709	0.2218	0.1126	151	-0.18	003.14	0
003	Fat, Crude (%)	2181	3.690	0.0800	3.709	0.2218	0.1126	151	-0.09	003.06	0
003	Fat, Crude (%)	0229	3.698	0.0050	3.709	0.2218	0.1126	151	-0.05	003.14	0
003	Fat, Crude (%)	0309	3.700	0.4000	3.709	0.2218	0.1126	151	-0.04	003.00	0
003	Fat, Crude (%)	0510	3.700	0.0000	3.709	0.2218	0.1126	151	-0.04	003.09	0
003	Fat, Crude (%)	0897	3.700	0.0200	3.709	0.2218	0.1126	151	-0.04	003.10	0
003	Fat, Crude (%)	2196	3.700	0.0000	3.709	0.2218	0.1126	151	-0.04	003.10	0
003	Fat, Crude (%)	0425	3.705	0.0100	3.709	0.2218	0.1126	151	-0.02	003.06	0
003	Fat, Crude (%)	2171	3.705	0.0300	3.709	0.2218	0.1126	151	-0.02	003.06	0
003	Fat, Crude (%)	2198	3.710	0.0200	3.709	0.2218	0.1126	151	0.00	003.00	0
003	Fat, Crude (%)	0019	3.710	0.3000	3.709	0.2218	0.1126	151	0.00	003.14	0
003	Fat, Crude (%)	0407	3.721	0.0120	3.709	0.2218	0.1126	151	0.05	003.14	0
003	Fat, Crude (%)	0541	3.725	0.3500	3.709	0.2218	0.1126	151	0.07	003.10	0
003	Fat, Crude (%)	0294	3.725	0.0500	3.709	0.2218	0.1126	151	0.07	003.06	0
003	Fat, Crude (%)	2268	3.725	0.4900	3.709	0.2218	0.1126	151	0.07	003.06	0
003	Fat, Crude (%)	0849	3.727	0.0070	3.709	0.2218	0.1126	151	0.08	003.06	0
003	Fat, Crude (%)	0878	3.730	0.0800	3.709	0.2218	0.1126	151	0.09	003.14	0
003	Fat, Crude (%)	2126	3.730	0.0200	3.709	0.2218	0.1126	151	0.09	003.00	0
003	Fat, Crude (%)	0175	3.740	0.2400	3.709	0.2218	0.1126	151	0.14	003.00	0
003	Fat, Crude (%)	0656	3.740	0.6000	3.709	0.2218	0.1126	151	0.14	003.05	0
003	Fat, Crude (%)	0169	3.740	0.0200	3.709	0.2218	0.1126	151	0.14	003.06	0
003	Fat, Crude (%)	0098	3.740	0.0400	3.709	0.2218	0.1126	151	0.14	003.10	0
003	Fat, Crude (%)	0682	3.740	0.0000	3.709	0.2218	0.1126	151	0.14	003.14	0
003	Fat, Crude (%)	0939	3.745	0.1500	3.709	0.2218	0.1126	151	0.16	003.06	0
003	Fat, Crude (%)	2006	3.745	0.0500	3.709	0.2218	0.1126	151	0.16	003.13	0
003	Fat, Crude (%)	0884	3.750	0.1000	3.709	0.2218	0.1126	151	0.18	003.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat, Crude (%)	0148	3.755	0.0100	3.709	0.2218	0.1126	151	0.21	003.14	0
003	Fat, Crude (%)	0955	3.760	0.0600	3.709	0.2218	0.1126	151	0.23	003.14	0
003	Fat, Crude (%)	0229	3.765	0.0060	3.709	0.2218	0.1126	151	0.25	003.06	0
003	Fat, Crude (%)	0028	3.770	0.2400	3.709	0.2218	0.1126	151	0.27	003.13	0
003	Fat, Crude (%)	2192	3.775	0.0500	3.709	0.2218	0.1126	151	0.30	003.00	0
003	Fat, Crude (%)	0242	3.775	0.0100	3.709	0.2218	0.1126	151	0.30	003.10	0
003	Fat, Crude (%)	0629	3.775	0.0300	3.709	0.2218	0.1126	151	0.30	003.10	0
003	Fat, Crude (%)	0027	3.783	0.1010	3.709	0.2218	0.1126	151	0.33	003.09	0
003	Fat, Crude (%)	0171	3.785	0.1700	3.709	0.2218	0.1126	151	0.34	003.12	0
003	Fat, Crude (%)	0001	3.794	0.1300	3.709	0.2218	0.1126	151	0.38	003.14	0
003	Fat, Crude (%)	2190	3.795	0.0300	3.709	0.2218	0.1126	151	0.39	003.00	0
003	Fat, Crude (%)	0675	3.800	0.0200	3.709	0.2218	0.1126	151	0.41	003.09	0
003	Fat, Crude (%)	0689	3.800	0.0000	3.709	0.2218	0.1126	151	0.41	003.10	0
003	Fat, Crude (%)	0529	3.800	0.0400	3.709	0.2218	0.1126	151	0.41	003.14	0
003	Fat, Crude (%)	0685	3.805	0.0500	3.709	0.2218	0.1126	151	0.43	003.09	0
003	Fat, Crude (%)	2109	3.805	0.2100	3.709	0.2218	0.1126	151	0.43	003.00	0
003	Fat, Crude (%)	0074	3.810	0.0400	3.709	0.2218	0.1126	151	0.45	003.14	0
003	Fat, Crude (%)	0619	3.810	0.1200	3.709	0.2218	0.1126	151	0.45	003.14	0
003	Fat, Crude (%)	0682	3.820	0.0000	3.709	0.2218	0.1126	151	0.50	003.06	0
003	Fat, Crude (%)	2076	3.821	0.0790	3.709	0.2218	0.1126	151	0.50	003.14	0
003	Fat, Crude (%)	0353	3.845	0.0100	3.709	0.2218	0.1126	151	0.61	003.09	0
003	Fat, Crude (%)	0187	3.845	0.0100	3.709	0.2218	0.1126	151	0.61	003.13	0
003	Fat, Crude (%)	2251	3.850	0.0600	3.709	0.2218	0.1126	151	0.63	003.07	0
003	Fat, Crude (%)	0910	3.850	0.1200	3.709	0.2218	0.1126	151	0.63	003.06	0
003	Fat, Crude (%)	0148	3.860	0.0000	3.709	0.2218	0.1126	151	0.68	003.10	0
003	Fat, Crude (%)	0870	3.862	0.2429	3.709	0.2218	0.1126	151	0.69	003.10	0
003	Fat, Crude (%)	0098	3.870	0.1600	3.709	0.2218	0.1126	151	0.72	003.13	0
003	Fat, Crude (%)	0098	3.890	0.0400	3.709	0.2218	0.1126	151	0.81	003.09	0
003	Fat, Crude (%)	0049	3.890	0.0000	3.709	0.2218	0.1126	151	0.81	003.14	0
003	Fat, Crude (%)	0689	3.900	0.0000	3.709	0.2218	0.1126	151	0.86	003.06	0
003	Fat, Crude (%)	0226	3.900	0.0000	3.709	0.2218	0.1126	151	0.86	003.09	0
003	Fat, Crude (%)	0695	3.915	0.0100	3.709	0.2218	0.1126	151	0.93	003.11	0
003	Fat, Crude (%)	0171	3.915	0.0300	3.709	0.2218	0.1126	151	0.93	003.14	0
003	Fat, Crude (%)	0644	3.920	0.0200	3.709	0.2218	0.1126	151	0.95	003.06	0
003	Fat, Crude (%)	2053	3.925	0.0300	3.709	0.2218	0.1126	151	0.97	003.00	0
003	Fat, Crude (%)	0618	3.929	0.0360	3.709	0.2218	0.1126	151	0.99	003.10	0
003	Fat, Crude (%)	0190	3.945	0.0100	3.709	0.2218	0.1126	151	1.06	003.14	0
003	Fat, Crude (%)	0692	3.950	0.3000	3.709	0.2218	0.1126	151	1.09	003.14	0
003	Fat, Crude (%)	0278	3.950	0.1000	3.709	0.2218	0.1126	151	1.09	003.14	0
003	Fat, Crude (%)	0004	3.965	0.4700	3.709	0.2218	0.1126	151	1.15	003.14	0
003	Fat, Crude (%)	0300	3.975	0.3100	3.709	0.2218	0.1126	151	1.20	003.06	0
003	Fat, Crude (%)	0726	3.975	0.0100	3.709	0.2218	0.1126	151	1.20	003.14	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat, Crude (%)	0687	4.000	0.4000	3.709	0.2218	0.1126	151	1.31	003.06	0
003	Fat, Crude (%)	0870	4.006	0.1630	3.709	0.2218	0.1126	151	1.34	003.09	0
003	Fat, Crude (%)	0553	4.010	0.0200	3.709	0.2218	0.1126	151	1.36	003.11	0
003	Fat, Crude (%)	0870	4.020	0.2117	3.709	0.2218	0.1126	151	1.40	003.13	0
003	Fat, Crude (%)	0017	4.050	0.1000	3.709	0.2218	0.1126	151	1.54	003.00	0
003	Fat, Crude (%)	2128	4.109	0.0513	3.709	0.2218	0.1126	151	1.80	003.11	0
003	Fat, Crude (%)	0590	4.145	0.0500	3.709	0.2218	0.1126	151	1.96	003.09	0
003	Fat, Crude (%)	0941	4.210	0.2200	3.709	0.2218	0.1126	151	2.26	003.99	0
003	Fat, Crude (%)	0164	4.255	0.0100	3.709	0.2218	0.1126	151	2.46	003.06	0
003	Fat, Crude (%)	0964	4.266	0.3460	3.709	0.2218	0.1126	151	2.51	003.09	0
003	Fat, Crude (%)	0676	4.290	0.0600	3.709	0.2218	0.1126	151	2.62	003.10	0
003	Fat, Crude (%)	0918	4.330	0.0400	3.709	0.2218	0.1126	151	2.80	003.06	0
003	Fat, Crude (%)	0265	4.405	0.0100	3.709	0.2218	0.1126	151	3.14	003.14	0
003	Fat, Crude (%)	2248	4.515	0.0100	3.709	0.2218	0.1126	151	3.63	003.01	0
003	Fat, Crude (%)	0011	4.600	0.0000	3.709	0.2218	0.1126	151	4.02	003.11	0
003	Fat, Crude (%)	0852	4.685	0.0100	3.709	0.2218	0.1126	151	4.40	003.11	0
003	Fat, Crude (%)	0297	4.825	0.0900	3.709	0.2218	0.1126	151	5.03	003.11	0
003	Fat, Crude (%)	0656	5.405	0.0300	3.709	0.2218	0.1126	151	7.65	003.06	0
003	Fat, Crude (%)	0889	2.960	2.200	3.709	0.2218	0.1126	151	-3.38	003.10	1
003	Fat, Crude (%)	0354	384.0	0.0000	3.709	0.2218	0.1126	151	1714.86	003.09	2
003	Fat, Crude (%)	0880	0.1000	0.2000	3.709	0.2218	0.1126	151		003.99	4
004	Fiber, Crude (%)	2199	2.350	0.1000	4.403	0.6876	0.2004	115	-2.99	004.06	0
004	Fiber, Crude (%)	2089	2.640	0.0200	4.403	0.6876	0.2004	115	-2.56	004.03	0
004	Fiber, Crude (%)	0042	2.845	0.0900	4.403	0.6876	0.2004	115	-2.27	004.07	0
004	Fiber, Crude (%)	0366	2.875	0.2900	4.403	0.6876	0.2004	115	-2.22	004.01	0
004	Fiber, Crude (%)	2248	3.040	0.0000	4.403	0.6876	0.2004	115	-1.98	004.99	0
004	Fiber, Crude (%)	0619	3.395	0.1900	4.403	0.6876	0.2004	115	-1.47	004.07	0
004	Fiber, Crude (%)	0629	3.425	0.0500	4.403	0.6876	0.2004	115	-1.42	004.99	0
004	Fiber, Crude (%)	0553	3.445	0.0900	4.403	0.6876	0.2004	115	-1.39	004.11	0
004	Fiber, Crude (%)	2033	3.450	0.3000	4.403	0.6876	0.2004	115	-1.39	004.07	0
004	Fiber, Crude (%)	0169	3.500	0.3000	4.403	0.6876	0.2004	115	-1.31	004.00	0
004	Fiber, Crude (%)	0682	3.600	0.0000	4.403	0.6876	0.2004	115	-1.17	004.07	0
004	Fiber, Crude (%)	0956	3.650	0.1000	4.403	0.6876	0.2004	115	-1.10	004.07	0
004	Fiber, Crude (%)	0868	3.730	0.1000	4.403	0.6876	0.2004	115	-0.98	004.06	0
004	Fiber, Crude (%)	0123	3.775	0.1100	4.403	0.6876	0.2004	115	-0.91	004.07	0
004	Fiber, Crude (%)	2190	3.805	0.4700	4.403	0.6876	0.2004	115	-0.87	004.07	0
004	Fiber, Crude (%)	2181	3.810	0.3200	4.403	0.6876	0.2004	115	-0.86	004.07	0
004	Fiber, Crude (%)	0885	3.838	0.0035	4.403	0.6876	0.2004	115	-0.82	004.06	0
004	Fiber, Crude (%)	0643	3.840	0.1200	4.403	0.6876	0.2004	115	-0.82	004.07	0
004	Fiber, Crude (%)	0425	3.850	0.1000	4.403	0.6876	0.2004	115	-0.80	004.00	0
004	Fiber, Crude (%)	0726	3.850	0.1000	4.403	0.6876	0.2004	115	-0.80	004.07	0
004	Fiber, Crude (%)	0870	3.866	0.0266	4.403	0.6876	0.2004	115	-0.78	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber, Crude (%)	0968	3.899	0.0460	4.403	0.6876	0.2004	115	-0.73	004.06	0
004	Fiber, Crude (%)	2022	3.910	0.0200	4.403	0.6876	0.2004	115	-0.72	004.06	0
004	Fiber, Crude (%)	0590	3.930	0.3400	4.403	0.6876	0.2004	115	-0.69	004.07	0
004	Fiber, Crude (%)	2109	3.930	0.0400	4.403	0.6876	0.2004	115	-0.69	004.07	0
004	Fiber, Crude (%)	0004	3.935	0.1100	4.403	0.6876	0.2004	115	-0.68	004.07	0
004	Fiber, Crude (%)	0644	3.944	0.5780	4.403	0.6876	0.2004	115	-0.67	004.07	0
004	Fiber, Crude (%)	2062	3.979	0.0325	4.403	0.6876	0.2004	115	-0.62	004.06	0
004	Fiber, Crude (%)	0164	4.000	0.2000	4.403	0.6876	0.2004	115	-0.59	004.00	0
004	Fiber, Crude (%)	0918	4.000	0.0400	4.403	0.6876	0.2004	115	-0.59	004.06	0
004	Fiber, Crude (%)	0011	4.000	0.0000	4.403	0.6876	0.2004	115	-0.59	004.11	0
004	Fiber, Crude (%)	0353	4.030	0.0200	4.403	0.6876	0.2004	115	-0.54	004.03	0
004	Fiber, Crude (%)	2006	4.040	0.0400	4.403	0.6876	0.2004	115	-0.53	004.06	0
004	Fiber, Crude (%)	0175	4.045	0.0900	4.403	0.6876	0.2004	115	-0.52	004.00	0
004	Fiber, Crude (%)	0100	4.045	0.5300	4.403	0.6876	0.2004	115	-0.52	004.07	0
004	Fiber, Crude (%)	0098	4.065	0.1300	4.403	0.6876	0.2004	115	-0.49	004.06	0
004	Fiber, Crude (%)	0354	4.065	0.0100	4.403	0.6876	0.2004	115	-0.49	004.07	0
004	Fiber, Crude (%)	0345	4.085	0.0500	4.403	0.6876	0.2004	115	-0.46	004.00	0
004	Fiber, Crude (%)	0884	4.100	0.4000	4.403	0.6876	0.2004	115	-0.44	004.00	0
004	Fiber, Crude (%)	0190	4.100	0.2000	4.403	0.6876	0.2004	115	-0.44	004.07	0
004	Fiber, Crude (%)	0610	4.100	0.0000	4.403	0.6876	0.2004	115	-0.44	004.99	0
004	Fiber, Crude (%)	0354	4.115	0.0300	4.403	0.6876	0.2004	115	-0.42	004.00	0
004	Fiber, Crude (%)	0968	4.115	0.1300	4.403	0.6876	0.2004	115	-0.42	004.07	0
004	Fiber, Crude (%)	2264	4.120	0.1846	4.403	0.6876	0.2004	115	-0.41	004.00	0
004	Fiber, Crude (%)	2193	4.120	0.0200	4.403	0.6876	0.2004	115	-0.41	004.07	0
004	Fiber, Crude (%)	2128	4.130	0.1075	4.403	0.6876	0.2004	115	-0.40	004.11	0
004	Fiber, Crude (%)	0083	4.145	0.1700	4.403	0.6876	0.2004	115	-0.38	004.07	0
004	Fiber, Crude (%)	2196	4.170	0.0000	4.403	0.6876	0.2004	115	-0.34	004.07	0
004	Fiber, Crude (%)	0144	4.190	0.0000	4.403	0.6876	0.2004	115	-0.31	004.07	0
004	Fiber, Crude (%)	0013	4.220	0.2000	4.403	0.6876	0.2004	115	-0.27	004.00	0
004	Fiber, Crude (%)	0309	4.221	0.0370	4.403	0.6876	0.2004	115	-0.27	004.00	0
004	Fiber, Crude (%)	0964	4.250	0.0800	4.403	0.6876	0.2004	115	-0.22	004.06	0
004	Fiber, Crude (%)	2073	4.260	0.1000	4.403	0.6876	0.2004	115	-0.21	004.00	0
004	Fiber, Crude (%)	2009	4.269	0.0463	4.403	0.6876	0.2004	115	-0.20	004.07	0
004	Fiber, Crude (%)	0720	4.270	0.1400	4.403	0.6876	0.2004	115	-0.19	004.06	0
004	Fiber, Crude (%)	0897	4.280	0.0000	4.403	0.6876	0.2004	115	-0.18	004.06	0
004	Fiber, Crude (%)	0265	4.295	0.3300	4.403	0.6876	0.2004	115	-0.16	004.07	0
004	Fiber, Crude (%)	0294	4.300	0.2000	4.403	0.6876	0.2004	115	-0.15	004.07	0
004	Fiber, Crude (%)	0171	4.310	0.0200	4.403	0.6876	0.2004	115	-0.14	004.00	0
004	Fiber, Crude (%)	2198	4.330	0.0400	4.403	0.6876	0.2004	115	-0.11	004.06	0
004	Fiber, Crude (%)	0910	4.350	0.1000	4.403	0.6876	0.2004	115	-0.08	004.07	0
004	Fiber, Crude (%)	0263	4.352	0.1160	4.403	0.6876	0.2004	115	-0.07	004.07	0
004	Fiber, Crude (%)	0019	4.355	0.2500	4.403	0.6876	0.2004	115	-0.07	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber, Crude (%)	0878	4.390	0.1800	4.403	0.6876	0.2004	115	-0.02	004.06	0
004	Fiber, Crude (%)	0505	4.395	0.0300	4.403	0.6876	0.2004	115	-0.01	004.07	0
004	Fiber, Crude (%)	0948	4.400	0.1000	4.403	0.6876	0.2004	115	0.00	004.06	0
004	Fiber, Crude (%)	0407	4.426	0.2645	4.403	0.6876	0.2004	115	0.03	004.07	0
004	Fiber, Crude (%)	0511	4.450	0.3000	4.403	0.6876	0.2004	115	0.07	004.07	0
004	Fiber, Crude (%)	0689	4.450	0.5000	4.403	0.6876	0.2004	115	0.07	004.06	0
004	Fiber, Crude (%)	2053	4.490	0.1800	4.403	0.6876	0.2004	115	0.13	004.07	0
004	Fiber, Crude (%)	0948	4.495	0.2300	4.403	0.6876	0.2004	115	0.13	004.07	0
004	Fiber, Crude (%)	0229	4.505	0.0100	4.403	0.6876	0.2004	115	0.15	004.07	0
004	Fiber, Crude (%)	2259	4.516	0.3154	4.403	0.6876	0.2004	115	0.16	004.07	0
004	Fiber, Crude (%)	0941	4.535	0.2700	4.403	0.6876	0.2004	115	0.19	004.06	0
004	Fiber, Crude (%)	2192	4.555	0.3300	4.403	0.6876	0.2004	115	0.22	004.03	0
004	Fiber, Crude (%)	0889	4.570	0.4600	4.403	0.6876	0.2004	115	0.24	004.11	0
004	Fiber, Crude (%)	0675	4.580	0.2000	4.403	0.6876	0.2004	115	0.26	004.07	0
004	Fiber, Crude (%)	0003	4.590	1.040	4.403	0.6876	0.2004	115	0.27	004.07	0
004	Fiber, Crude (%)	2076	4.593	0.6210	4.403	0.6876	0.2004	115	0.28	004.00	0
004	Fiber, Crude (%)	0298	4.600	0.2000	4.403	0.6876	0.2004	115	0.29	004.07	0
004	Fiber, Crude (%)	0881	4.604	0.0260	4.403	0.6876	0.2004	115	0.29	004.06	0
004	Fiber, Crude (%)	0708	4.660	0.2000	4.403	0.6876	0.2004	115	0.37	004.07	0
004	Fiber, Crude (%)	0510	4.700	0.0000	4.403	0.6876	0.2004	115	0.43	004.07	0
004	Fiber, Crude (%)	0581	4.700	1.000	4.403	0.6876	0.2004	115	0.43	004.07	0
004	Fiber, Crude (%)	0098	4.720	0.8000	4.403	0.6876	0.2004	115	0.46	004.07	0
004	Fiber, Crude (%)	0520	4.785	0.0100	4.403	0.6876	0.2004	115	0.56	004.07	0
004	Fiber, Crude (%)	0226	4.800	0.0000	4.403	0.6876	0.2004	115	0.58	004.00	0
004	Fiber, Crude (%)	0066	4.807	0.1030	4.403	0.6876	0.2004	115	0.59	004.07	0
004	Fiber, Crude (%)	0297	4.855	1.090	4.403	0.6876	0.2004	115	0.66	004.07	0
004	Fiber, Crude (%)	0278	4.900	0.2000	4.403	0.6876	0.2004	115	0.72	004.07	0
004	Fiber, Crude (%)	0675	4.910	0.2000	4.403	0.6876	0.2004	115	0.74	004.06	0
004	Fiber, Crude (%)	0683	4.960	0.2000	4.403	0.6876	0.2004	115	0.81	004.07	0
004	Fiber, Crude (%)	0504	5.035	0.0900	4.403	0.6876	0.2004	115	0.92	004.00	0
004	Fiber, Crude (%)	2246	5.130	0.1400	4.403	0.6876	0.2004	115	1.06	004.03	0
004	Fiber, Crude (%)	0536	5.180	0.0200	4.403	0.6876	0.2004	115	1.13	004.07	0
004	Fiber, Crude (%)	0893	5.200	0.0000	4.403	0.6876	0.2004	115	1.16	004.06	0
004	Fiber, Crude (%)	0692	5.200	0.6000	4.403	0.6876	0.2004	115	1.16	004.07	0
004	Fiber, Crude (%)	0034	5.325	0.7300	4.403	0.6876	0.2004	115	1.34	004.07	0
004	Fiber, Crude (%)	0089	5.395	0.0100	4.403	0.6876	0.2004	115	1.44	004.07	0
004	Fiber, Crude (%)	0074	5.405	0.4900	4.403	0.6876	0.2004	115	1.46	004.07	0
004	Fiber, Crude (%)	0045	5.455	0.8900	4.403	0.6876	0.2004	115	1.53	004.07	0
004	Fiber, Crude (%)	0529	5.455	0.0300	4.403	0.6876	0.2004	115	1.53	004.07	0
004	Fiber, Crude (%)	0876	5.475	0.7700	4.403	0.6876	0.2004	115	1.56	004.07	0
004	Fiber, Crude (%)	0242	5.515	0.2500	4.403	0.6876	0.2004	115	1.62	004.07	0
004	Fiber, Crude (%)	0032	5.530	0.0200	4.403	0.6876	0.2004	115	1.64	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber, Crude (%)	0676	5.535	0.2100	4.403	0.6876	0.2004	115	1.65	004.06	0
004	Fiber, Crude (%)	0848	5.580	0.0000	4.403	0.6876	0.2004	115	1.71	004.07	0
004	Fiber, Crude (%)	0720	5.600	0.0000	4.403	0.6876	0.2004	115	1.74	004.11	0
004	Fiber, Crude (%)	0852	5.620	0.0200	4.403	0.6876	0.2004	115	1.77	004.11	0
004	Fiber, Crude (%)	0939	6.260	0.4200	4.403	0.6876	0.2004	115	2.70	004.07	0
004	Fiber, Crude (%)	0656	6.380	0.0000	4.403	0.6876	0.2004	115	2.87	004.07	0
004	Fiber, Crude (%)	0849	6.891	0.0420	4.403	0.6876	0.2004	115	3.62	004.07	0
004	Fiber, Crude (%)	0626	8.025	0.1900	4.403	0.6876	0.2004	115	5.27	004.07	0
004	Fiber, Crude (%)	0693	8.341	1.004	4.403	0.6876	0.2004	115	5.73	004.07	0
004	Fiber, Crude (%)	0955	10.40	0.0000	4.403	0.6876	0.2004	115	8.72	004.07	0
004	Fiber, Crude (%)	0202	5.644	1.933	4.403	0.6876	0.2004	115	1.80	004.07	1
004	Fiber, Crude (%)	0598	9.240	2.920	4.403	0.6876	0.2004	115	7.03	004.07	1
005	Ash (%)	0889	9.310	0.1400	15.17	0.5856	0.1736	145	-10.01	005.11	0
005	Ash (%)	0297	9.550	0.1000	15.17	0.5856	0.1736	145	-9.60	005.11	0
005	Ash (%)	2128	9.602	0.5600	15.17	0.5856	0.1736	145	-9.51	005.11	0
005	Ash (%)	0852	11.22	0.0800	15.17	0.5856	0.1736	145	-6.74	005.11	0
005	Ash (%)	2081	11.39	0.3400	15.17	0.5856	0.1736	145	-6.45	005.99	0
005	Ash (%)	2184	11.67	0.0200	15.17	0.5856	0.1736	145	-5.98	005.00	0
005	Ash (%)	0720	11.84	0.1200	15.17	0.5856	0.1736	145	-5.69	005.11	0
005	Ash (%)	0169	12.05	0.0100	15.17	0.5856	0.1736	145	-5.34	005.00	0
005	Ash (%)	0539	12.21	0.3100	15.17	0.5856	0.1736	145	-5.06	005.00	0
005	Ash (%)	0662	12.43	0.0848	15.17	0.5856	0.1736	145	-4.67	005.99	0
005	Ash (%)	2033	12.70	0.2000	15.17	0.5856	0.1736	145	-4.22	005.00	0
005	Ash (%)	0650	13.05	0.9000	15.17	0.5856	0.1736	145	-3.62	005.00	0
005	Ash (%)	0278	13.20	0.9200	15.17	0.5856	0.1736	145	-3.36	005.00	0
005	Ash (%)	0049	13.24	1.560	15.17	0.5856	0.1736	145	-3.30	005.00	0
005	Ash (%)	0511	13.36	0.5300	15.17	0.5856	0.1736	145	-3.10	005.00	0
005	Ash (%)	0357	13.45	0.1000	15.17	0.5856	0.1736	145	-2.94	005.00	0
005	Ash (%)	0598	13.83	0.0900	15.17	0.5856	0.1736	145	-2.30	005.00	0
005	Ash (%)	0148	13.92	0.0200	15.17	0.5856	0.1736	145	-2.13	005.00	0
005	Ash (%)	0425	13.95	0.0200	15.17	0.5856	0.1736	145	-2.08	005.00	0
005	Ash (%)	0918	14.17	0.1700	15.17	0.5856	0.1736	145	-1.72	005.00	0
005	Ash (%)	2073	14.18	0.0500	15.17	0.5856	0.1736	145	-1.70	005.00	0
005	Ash (%)	0574	14.24	0.1300	15.17	0.5856	0.1736	145	-1.60	005.00	0
005	Ash (%)	0019	14.30	0.2000	15.17	0.5856	0.1736	145	-1.49	005.00	0
005	Ash (%)	0003	14.32	0.4500	15.17	0.5856	0.1736	145	-1.46	005.05	0
005	Ash (%)	0912	14.35	0.1000	15.17	0.5856	0.1736	145	-1.40	005.03	0
005	Ash (%)	2076	14.39	0.5995	15.17	0.5856	0.1736	145	-1.33	005.00	0
005	Ash (%)	0693	14.47	0.2260	15.17	0.5856	0.1736	145	-1.19	005.00	0
005	Ash (%)	0226	14.55	0.5000	15.17	0.5856	0.1736	145	-1.06	005.00	0
005	Ash (%)	0229	14.56	0.3100	15.17	0.5856	0.1736	145	-1.05	005.00	0
005	Ash (%)	0682	14.58	0.0000	15.17	0.5856	0.1736	145	-1.01	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0876	14.60	0.2500	15.17	0.5856	0.1736	145	-0.98	005.00	0
005	Ash (%)	2022	14.65	0.0100	15.17	0.5856	0.1736	145	-0.90	005.00	0
005	Ash (%)	0004	14.67	0.0400	15.17	0.5856	0.1736	145	-0.85	005.00	0
005	Ash (%)	0726	14.67	0.0400	15.17	0.5856	0.1736	145	-0.85	005.05	0
005	Ash (%)	0720	14.68	0.0100	15.17	0.5856	0.1736	145	-0.84	005.05	0
005	Ash (%)	0644	14.74	0.1400	15.17	0.5856	0.1736	145	-0.73	005.02	0
005	Ash (%)	0298	14.75	0.0200	15.17	0.5856	0.1736	145	-0.72	005.00	0
005	Ash (%)	0066	14.83	0.1230	15.17	0.5856	0.1736	145	-0.59	005.05	0
005	Ash (%)	0358	14.84	0.1500	15.17	0.5856	0.1736	145	-0.57	005.00	0
005	Ash (%)	0541	14.84	0.6700	15.17	0.5856	0.1736	145	-0.57	005.00	0
005	Ash (%)	0042	14.89	0.0500	15.17	0.5856	0.1736	145	-0.49	005.00	0
005	Ash (%)	2053	14.94	0.2900	15.17	0.5856	0.1736	145	-0.40	005.00	0
005	Ash (%)	0038	14.96	0.3200	15.17	0.5856	0.1736	145	-0.36	005.05	0
005	Ash (%)	0026	15.00	0.3900	15.17	0.5856	0.1736	145	-0.30	005.00	0
005	Ash (%)	2199	15.00	0.4000	15.17	0.5856	0.1736	145	-0.29	005.00	0
005	Ash (%)	0881	15.00	0.1070	15.17	0.5856	0.1736	145	-0.29	005.00	0
005	Ash (%)	2150	15.02	0.2300	15.17	0.5856	0.1736	145	-0.26	005.00	0
005	Ash (%)	0297	15.04	0.2800	15.17	0.5856	0.1736	145	-0.22	005.00	0
005	Ash (%)	0853	15.05	0.1000	15.17	0.5856	0.1736	145	-0.20	005.00	0
005	Ash (%)	2188	15.08	0.0600	15.17	0.5856	0.1736	145	-0.15	005.00	0
005	Ash (%)	0142	15.10	0.0000	15.17	0.5856	0.1736	145	-0.12	005.00	0
005	Ash (%)	0953	15.11	0.2700	15.17	0.5856	0.1736	145	-0.11	005.00	0
005	Ash (%)	0083	15.14	0.0100	15.17	0.5856	0.1736	145	-0.06	005.00	0
005	Ash (%)	0504	15.17	0.2400	15.17	0.5856	0.1736	145	0.00	005.00	0
005	Ash (%)	0175	15.18	0.3000	15.17	0.5856	0.1736	145	0.02	005.00	0
005	Ash (%)	0897	15.19	0.0200	15.17	0.5856	0.1736	145	0.03	005.05	0
005	Ash (%)	0098	15.19	0.0800	15.17	0.5856	0.1736	145	0.03	005.00	0
005	Ash (%)	0968	15.20	0.0400	15.17	0.5856	0.1736	145	0.05	005.00	0
005	Ash (%)	0510	15.23	0.0300	15.17	0.5856	0.1736	145	0.09	005.00	0
005	Ash (%)	2259	15.24	0.1870	15.17	0.5856	0.1736	145	0.12	005.00	0
005	Ash (%)	0870	15.25	0.0032	15.17	0.5856	0.1736	145	0.13	005.00	0
005	Ash (%)	0033	15.25	0.1000	15.17	0.5856	0.1736	145	0.14	005.00	0
005	Ash (%)	2242	15.27	0.3100	15.17	0.5856	0.1736	145	0.16	005.00	0
005	Ash (%)	0529	15.27	0.5600	15.17	0.5856	0.1736	145	0.17	005.00	0
005	Ash (%)	2248	15.29	0.0100	15.17	0.5856	0.1736	145	0.20	005.00	0
005	Ash (%)	0618	15.29	0.1157	15.17	0.5856	0.1736	145	0.20	005.00	0
005	Ash (%)	0520	15.30	0.0100	15.17	0.5856	0.1736	145	0.21	005.00	0
005	Ash (%)	2109	15.30	0.3600	15.17	0.5856	0.1736	145	0.22	005.05	0
005	Ash (%)	0692	15.31	0.2800	15.17	0.5856	0.1736	145	0.24	005.05	0
005	Ash (%)	0100	15.32	0.1700	15.17	0.5856	0.1736	145	0.25	005.00	0
005	Ash (%)	0848	15.32	0.0000	15.17	0.5856	0.1736	145	0.26	005.00	0
005	Ash (%)	0656	15.33	0.1200	15.17	0.5856	0.1736	145	0.27	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0035	15.34	0.0100	15.17	0.5856	0.1736	145	0.28	005.00	0
005	Ash (%)	0001	15.35	0.0978	15.17	0.5856	0.1736	145	0.31	005.00	0
005	Ash (%)	0265	15.35	0.1000	15.17	0.5856	0.1736	145	0.31	005.00	0
005	Ash (%)	0619	15.35	0.1000	15.17	0.5856	0.1736	145	0.31	005.05	0
005	Ash (%)	2196	15.36	0.0000	15.17	0.5856	0.1736	145	0.32	005.00	0
005	Ash (%)	0643	15.38	0.0300	15.17	0.5856	0.1736	145	0.35	005.00	0
005	Ash (%)	0880	15.38	0.2500	15.17	0.5856	0.1736	145	0.35	005.00	0
005	Ash (%)	0047	15.38	0.1800	15.17	0.5856	0.1736	145	0.36	005.00	0
005	Ash (%)	0108	15.38	0.0200	15.17	0.5856	0.1736	145	0.36	005.00	0
005	Ash (%)	0874	15.38	0.1600	15.17	0.5856	0.1736	145	0.36	005.00	0
005	Ash (%)	0242	15.39	0.1800	15.17	0.5856	0.1736	145	0.38	005.00	0
005	Ash (%)	2246	15.40	0.0700	15.17	0.5856	0.1736	145	0.38	005.00	0
005	Ash (%)	0884	15.40	0.6000	15.17	0.5856	0.1736	145	0.39	005.99	0
005	Ash (%)	0144	15.40	0.0000	15.17	0.5856	0.1736	145	0.39	005.00	0
005	Ash (%)	0889	15.41	0.0900	15.17	0.5856	0.1736	145	0.40	005.05	0
005	Ash (%)	0505	15.42	0.0500	15.17	0.5856	0.1736	145	0.42	005.00	0
005	Ash (%)	2089	15.43	0.1100	15.17	0.5856	0.1736	145	0.44	005.00	0
005	Ash (%)	0354	15.43	0.0600	15.17	0.5856	0.1736	145	0.44	005.00	0
005	Ash (%)	0407	15.44	0.0725	15.17	0.5856	0.1736	145	0.45	005.00	0
005	Ash (%)	0589	15.45	0.1100	15.17	0.5856	0.1736	145	0.47	005.00	0
005	Ash (%)	0629	15.45	0.0700	15.17	0.5856	0.1736	145	0.47	005.99	0
005	Ash (%)	2193	15.45	0.0200	15.17	0.5856	0.1736	145	0.48	005.00	0
005	Ash (%)	2006	15.46	0.0200	15.17	0.5856	0.1736	145	0.50	005.00	0
005	Ash (%)	0034	15.47	0.2760	15.17	0.5856	0.1736	145	0.50	005.00	0
005	Ash (%)	0596	15.48	0.0100	15.17	0.5856	0.1736	145	0.52	005.05	0
005	Ash (%)	0187	15.48	0.0400	15.17	0.5856	0.1736	145	0.53	005.00	0
005	Ash (%)	0263	15.49	0.0580	15.17	0.5856	0.1736	145	0.54	005.05	0
005	Ash (%)	0045	15.50	0.4000	15.17	0.5856	0.1736	145	0.56	005.00	0
005	Ash (%)	0956	15.50	0.8000	15.17	0.5856	0.1736	145	0.56	005.00	0
005	Ash (%)	0878	15.50	0.0800	15.17	0.5856	0.1736	145	0.56	005.05	0
005	Ash (%)	2268	15.51	0.1400	15.17	0.5856	0.1736	145	0.58	005.00	0
005	Ash (%)	2198	15.52	0.0500	15.17	0.5856	0.1736	145	0.59	005.05	0
005	Ash (%)	0164	15.52	0.1400	15.17	0.5856	0.1736	145	0.60	005.00	0
005	Ash (%)	2181	15.52	0.0800	15.17	0.5856	0.1736	145	0.60	005.00	0
005	Ash (%)	0689	15.55	0.0500	15.17	0.5856	0.1736	145	0.64	005.05	0
005	Ash (%)	0017	15.55	0.1000	15.17	0.5856	0.1736	145	0.65	005.00	0
005	Ash (%)	0171	15.55	0.3000	15.17	0.5856	0.1736	145	0.65	005.00	0
005	Ash (%)	0309	15.55	0.1400	15.17	0.5856	0.1736	145	0.65	005.00	0
005	Ash (%)	2192	15.56	0.0500	15.17	0.5856	0.1736	145	0.66	005.05	0
005	Ash (%)	2264	15.56	0.0870	15.17	0.5856	0.1736	145	0.66	005.05	0
005	Ash (%)	0849	15.56	0.0320	15.17	0.5856	0.1736	145	0.67	005.05	0
005	Ash (%)	0695	15.57	0.0100	15.17	0.5856	0.1736	145	0.67	005.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0202	15.60	0.1180	15.17	0.5856	0.1736	145	0.73	005.99	0
005	Ash (%)	0675	15.60	0.2600	15.17	0.5856	0.1736	145	0.73	005.00	0
005	Ash (%)	0683	15.61	0.0300	15.17	0.5856	0.1736	145	0.74	005.05	0
005	Ash (%)	0366	15.65	0.1000	15.17	0.5856	0.1736	145	0.82	005.00	0
005	Ash (%)	0294	15.65	0.1000	15.17	0.5856	0.1736	145	0.82	005.05	0
005	Ash (%)	0590	15.65	0.0600	15.17	0.5856	0.1736	145	0.82	005.05	0
005	Ash (%)	0536	15.65	0.1000	15.17	0.5856	0.1736	145	0.82	005.99	0
005	Ash (%)	0910	15.65	0.5000	15.17	0.5856	0.1736	145	0.82	005.99	0
005	Ash (%)	0062	15.67	0.0980	15.17	0.5856	0.1736	145	0.86	005.00	0
005	Ash (%)	2009	15.68	0.0084	15.17	0.5856	0.1736	145	0.87	005.05	0
005	Ash (%)	0676	15.69	0.0900	15.17	0.5856	0.1736	145	0.88	005.99	0
005	Ash (%)	2246	15.69	0.1200	15.17	0.5856	0.1736	145	0.89	005.05	0
005	Ash (%)	0345	15.70	0.2000	15.17	0.5856	0.1736	145	0.91	005.00	0
005	Ash (%)	0581	15.70	0.0200	15.17	0.5856	0.1736	145	0.91	005.00	0
005	Ash (%)	2126	15.70	0.0400	15.17	0.5856	0.1736	145	0.91	005.05	0
005	Ash (%)	0941	15.72	0.2000	15.17	0.5856	0.1736	145	0.94	005.05	0
005	Ash (%)	0939	15.73	0.0100	15.17	0.5856	0.1736	145	0.95	005.05	0
005	Ash (%)	0652	15.75	0.1000	15.17	0.5856	0.1736	145	0.99	005.99	0
005	Ash (%)	2062	15.75	0.0126	15.17	0.5856	0.1736	145	0.99	005.05	0
005	Ash (%)	2190	15.77	0.2200	15.17	0.5856	0.1736	145	1.02	005.05	0
005	Ash (%)	2251	15.77	0.0200	15.17	0.5856	0.1736	145	1.02	005.99	0
005	Ash (%)	0685	15.78	0.0000	15.17	0.5856	0.1736	145	1.04	005.05	0
005	Ash (%)	0353	15.78	0.1000	15.17	0.5856	0.1736	145	1.04	005.00	0
005	Ash (%)	0885	15.80	0.0789	15.17	0.5856	0.1736	145	1.08	005.05	0
005	Ash (%)	0123	15.81	0.3500	15.17	0.5856	0.1736	145	1.08	005.00	0
005	Ash (%)	0190	15.82	0.1200	15.17	0.5856	0.1736	145	1.11	005.05	0
005	Ash (%)	0868	15.83	0.0000	15.17	0.5856	0.1736	145	1.13	005.05	0
005	Ash (%)	0852	15.99	0.0700	15.17	0.5856	0.1736	145	1.39	005.00	0
005	Ash (%)	0948	15.99	0.0300	15.17	0.5856	0.1736	145	1.39	005.00	0
005	Ash (%)	0610	16.13	0.9400	15.17	0.5856	0.1736	145	1.64	005.05	0
005	Ash (%)	2129	16.32	0.1000	15.17	0.5856	0.1736	145	1.96	005.05	0
005	Ash (%)	0893	755.0	1,480	15.17	0.5856	0.1736	145	1263.25	005.05	2
006	Total Sugars (%)	0407	2.010	0.8600	3.107	0.8509	0.4060	5		006.00	0
006	Total Sugars (%)	0148	2.465	0.1900	3.107	0.8509	0.4060	5		006.99	0
006	Total Sugars (%)	0910	3.390	0.1000	3.107	0.8509	0.4060	5		006.00	0
006	Total Sugars (%)	0720	3.570	0.2800	3.107	0.8509	0.4060	5		006.99	0
006	Total Sugars (%)	0956	4.100	0.6000	3.107	0.8509	0.4060	5		006.99	0
008	Fiber, Acid Detergent (%)	0026	4.315	0.6300	5.692	0.5613	0.2587	61	-2.45	008.08	0
008	Fiber, Acid Detergent (%)	0878	4.600	0.1600	5.692	0.5613	0.2587	61	-1.95	008.08	0
008	Fiber, Acid Detergent (%)	0910	4.795	0.3500	5.692	0.5613	0.2587	61	-1.60	008.08	0
008	Fiber, Acid Detergent (%)	0590	4.800	0.4400	5.692	0.5613	0.2587	61	-1.59	008.08	0
008	Fiber, Acid Detergent (%)	0939	4.800	0.1200	5.692	0.5613	0.2587	61	-1.59	008.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
008	Fiber, Acid Detergent (%)	0676	4.805	0.1100	5.692	0.5613	0.2587	61	-1.58	008.02	0
008	Fiber, Acid Detergent (%)	0956	5.050	0.3000	5.692	0.5613	0.2587	61	-1.14	008.08	0
008	Fiber, Acid Detergent (%)	0004	5.060	0.1200	5.692	0.5613	0.2587	61	-1.13	008.08	0
008	Fiber, Acid Detergent (%)	0148	5.170	0.5200	5.692	0.5613	0.2587	61	-0.93	008.08	0
008	Fiber, Acid Detergent (%)	0297	5.170	0.6600	5.692	0.5613	0.2587	61	-0.93	008.08	0
008	Fiber, Acid Detergent (%)	0510	5.200	0.0000	5.692	0.5613	0.2587	61	-0.88	008.08	0
008	Fiber, Acid Detergent (%)	0504	5.340	0.2800	5.692	0.5613	0.2587	61	-0.63	008.02	0
008	Fiber, Acid Detergent (%)	2193	5.340	0.3200	5.692	0.5613	0.2587	61	-0.63	008.08	0
008	Fiber, Acid Detergent (%)	0049	5.360	0.1400	5.692	0.5613	0.2587	61	-0.59	008.08	0
008	Fiber, Acid Detergent (%)	0148	5.375	0.7100	5.692	0.5613	0.2587	61	-0.56	008.02	0
008	Fiber, Acid Detergent (%)	2246	5.380	0.0200	5.692	0.5613	0.2587	61	-0.56	008.02	0
008	Fiber, Acid Detergent (%)	0948	5.390	0.3200	5.692	0.5613	0.2587	61	-0.54	008.08	0
008	Fiber, Acid Detergent (%)	0263	5.402	0.0010	5.692	0.5613	0.2587	61	-0.52	008.08	0
008	Fiber, Acid Detergent (%)	0035	5.450	0.0200	5.692	0.5613	0.2587	61	-0.43	008.08	0
008	Fiber, Acid Detergent (%)	0164	5.450	0.1000	5.692	0.5613	0.2587	61	-0.43	008.08	0
008	Fiber, Acid Detergent (%)	2196	5.460	0.0000	5.692	0.5613	0.2587	61	-0.41	008.08	0
008	Fiber, Acid Detergent (%)	0848	5.485	0.0100	5.692	0.5613	0.2587	61	-0.37	008.08	0
008	Fiber, Acid Detergent (%)	0278	5.550	0.3000	5.692	0.5613	0.2587	61	-0.25	008.08	0
008	Fiber, Acid Detergent (%)	0581	5.555	0.4100	5.692	0.5613	0.2587	61	-0.24	008.08	0
008	Fiber, Acid Detergent (%)	2129	5.565	0.5500	5.692	0.5613	0.2587	61	-0.23	008.08	0
008	Fiber, Acid Detergent (%)	0309	5.572	0.0715	5.692	0.5613	0.2587	61	-0.21	008.02	0
008	Fiber, Acid Detergent (%)	0098	5.585	0.6100	5.692	0.5613	0.2587	61	-0.19	008.02	0
008	Fiber, Acid Detergent (%)	0948	5.590	0.0800	5.692	0.5613	0.2587	61	-0.18	008.02	0
008	Fiber, Acid Detergent (%)	0511	5.600	0.4000	5.692	0.5613	0.2587	61	-0.16	008.99	0
008	Fiber, Acid Detergent (%)	2192	5.610	0.7000	5.692	0.5613	0.2587	61	-0.15	008.02	0
008	Fiber, Acid Detergent (%)	0066	5.632	0.4210	5.692	0.5613	0.2587	61	-0.11	008.08	0
008	Fiber, Acid Detergent (%)	0083	5.650	0.2800	5.692	0.5613	0.2587	61	-0.07	008.08	0
008	Fiber, Acid Detergent (%)	0042	5.670	0.2600	5.692	0.5613	0.2587	61	-0.04	008.08	0
008	Fiber, Acid Detergent (%)	0357	5.700	0.0000	5.692	0.5613	0.2587	61	0.01	008.08	0
008	Fiber, Acid Detergent (%)	0037	5.720	0.3600	5.692	0.5613	0.2587	61	0.05	008.08	0
008	Fiber, Acid Detergent (%)	0675	5.725	0.2900	5.692	0.5613	0.2587	61	0.06	008.08	0
008	Fiber, Acid Detergent (%)	0187	5.750	0.0800	5.692	0.5613	0.2587	61	0.10	008.02	0
008	Fiber, Acid Detergent (%)	0407	5.780	0.0476	5.692	0.5613	0.2587	61	0.16	008.08	0
008	Fiber, Acid Detergent (%)	0345	5.855	0.1100	5.692	0.5613	0.2587	61	0.29	008.02	0
008	Fiber, Acid Detergent (%)	0870	5.868	0.2588	5.692	0.5613	0.2587	61	0.31	008.08	0
008	Fiber, Acid Detergent (%)	0032	5.875	0.1700	5.692	0.5613	0.2587	61	0.33	008.08	0
008	Fiber, Acid Detergent (%)	0353	5.905	0.0300	5.692	0.5613	0.2587	61	0.38	008.02	0
008	Fiber, Acid Detergent (%)	0358	5.930	0.1400	5.692	0.5613	0.2587	61	0.42	008.08	0
008	Fiber, Acid Detergent (%)	0226	5.950	0.3000	5.692	0.5613	0.2587	61	0.46	008.02	0
008	Fiber, Acid Detergent (%)	0354	5.985	0.0500	5.692	0.5613	0.2587	61	0.52	008.08	0
008	Fiber, Acid Detergent (%)	0675	5.990	0.0800	5.692	0.5613	0.2587	61	0.53	008.02	0
008	Fiber, Acid Detergent (%)	0610	6.000	0.2000	5.692	0.5613	0.2587	61	0.55	008.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
008	Fiber, Acid Detergent (%)	0001	6.025	0.1900	5.692	0.5613	0.2587	61	0.59	008.08	0
008	Fiber, Acid Detergent (%)	0294	6.050	0.1000	5.692	0.5613	0.2587	61	0.64	008.08	0
008	Fiber, Acid Detergent (%)	0964	6.155	0.1930	5.692	0.5613	0.2587	61	0.82	008.02	0
008	Fiber, Acid Detergent (%)	0190	6.250	0.1000	5.692	0.5613	0.2587	61	0.99	008.08	0
008	Fiber, Acid Detergent (%)	0726	6.300	0.0800	5.692	0.5613	0.2587	61	1.08	008.08	0
008	Fiber, Acid Detergent (%)	0720	6.420	0.0800	5.692	0.5613	0.2587	61	1.30	008.99	0
008	Fiber, Acid Detergent (%)	0918	6.481	0.8022	5.692	0.5613	0.2587	61	1.40	008.08	0
008	Fiber, Acid Detergent (%)	0045	6.550	0.9800	5.692	0.5613	0.2587	61	1.53	008.08	0
008	Fiber, Acid Detergent (%)	0693	6.842	0.4740	5.692	0.5613	0.2587	61	2.05	008.08	0
008	Fiber, Acid Detergent (%)	0265	7.090	0.5000	5.692	0.5613	0.2587	61	2.49	008.05	0
008	Fiber, Acid Detergent (%)	0968	7.366	0.1210	5.692	0.5613	0.2587	61	2.98	008.08	0
008	Fiber, Acid Detergent (%)	0889	7.440	0.0600	5.692	0.5613	0.2587	61	3.11	008.08	0
008	Fiber, Acid Detergent (%)	0536	7.475	0.3700	5.692	0.5613	0.2587	61	3.18	008.08	0
008	Fiber, Acid Detergent (%)	0619	12.40	0.2000	5.692	0.5613	0.2587	61	11.95	008.08	0
008	Fiber, Acid Detergent (%)	0656	21.36	1.010	5.692	0.5613	0.2587	61	27.91	008.05	2
009	Fiber, Neutral Detergent (%)	0619	5.080	0.3400	14.85	1.461	0.3690	58	-6.69	009.09	0
009	Fiber, Neutral Detergent (%)	0598	12.53	0.1300	14.85	1.461	0.3690	58	-1.59	009.09	0
009	Fiber, Neutral Detergent (%)	0720	12.61	0.0500	14.85	1.461	0.3690	58	-1.54	009.99	0
009	Fiber, Neutral Detergent (%)	0049	12.69	0.1900	14.85	1.461	0.3690	58	-1.48	009.09	0
009	Fiber, Neutral Detergent (%)	0910	13.00	0.2000	14.85	1.461	0.3690	58	-1.27	009.07	0
009	Fiber, Neutral Detergent (%)	0939	13.28	0.2900	14.85	1.461	0.3690	58	-1.08	009.09	0
009	Fiber, Neutral Detergent (%)	0164	13.30	0.4000	14.85	1.461	0.3690	58	-1.06	009.09	0
009	Fiber, Neutral Detergent (%)	0581	13.32	0.4100	14.85	1.461	0.3690	58	-1.05	009.09	0
009	Fiber, Neutral Detergent (%)	0878	13.40	0.2000	14.85	1.461	0.3690	58	-1.00	009.09	0
009	Fiber, Neutral Detergent (%)	0263	13.43	0.0220	14.85	1.461	0.3690	58	-0.97	009.09	0
009	Fiber, Neutral Detergent (%)	0187	13.54	0.0600	14.85	1.461	0.3690	58	-0.90	009.07	0
009	Fiber, Neutral Detergent (%)	0357	13.70	0.2000	14.85	1.461	0.3690	58	-0.79	009.09	0
009	Fiber, Neutral Detergent (%)	0297	13.71	0.8600	14.85	1.461	0.3690	58	-0.78	009.09	0
009	Fiber, Neutral Detergent (%)	0956	13.90	0.4000	14.85	1.461	0.3690	58	-0.65	009.09	0
009	Fiber, Neutral Detergent (%)	0345	13.95	0.3000	14.85	1.461	0.3690	58	-0.62	009.07	0
009	Fiber, Neutral Detergent (%)	0504	13.98	0.0600	14.85	1.461	0.3690	58	-0.60	009.04	0
009	Fiber, Neutral Detergent (%)	0353	13.99	0.1500	14.85	1.461	0.3690	58	-0.60	009.07	0
009	Fiber, Neutral Detergent (%)	0407	14.03	0.5945	14.85	1.461	0.3690	58	-0.57	009.09	0
009	Fiber, Neutral Detergent (%)	0098	14.05	0.3000	14.85	1.461	0.3690	58	-0.55	009.07	0
009	Fiber, Neutral Detergent (%)	0294	14.10	0.2000	14.85	1.461	0.3690	58	-0.52	009.09	0
009	Fiber, Neutral Detergent (%)	2193	14.20	0.0600	14.85	1.461	0.3690	58	-0.45	009.09	0
009	Fiber, Neutral Detergent (%)	0889	14.30	0.0700	14.85	1.461	0.3690	58	-0.38	009.99	0
009	Fiber, Neutral Detergent (%)	0948	14.32	0.2100	14.85	1.461	0.3690	58	-0.37	009.09	0
009	Fiber, Neutral Detergent (%)	0001	14.37	0.0600	14.85	1.461	0.3690	58	-0.33	009.09	0
009	Fiber, Neutral Detergent (%)	0948	14.38	0.4300	14.85	1.461	0.3690	58	-0.33	009.07	0
009	Fiber, Neutral Detergent (%)	0358	14.52	0.3000	14.85	1.461	0.3690	58	-0.23	009.09	0
009	Fiber, Neutral Detergent (%)	2196	14.62	0.0000	14.85	1.461	0.3690	58	-0.16	009.09	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
009	Fiber, Neutral Detergent (%)	0148	14.66	1.080	14.85	1.461	0.3690	58	-0.13	009.09	0
009	Fiber, Neutral Detergent (%)	0354	14.69	0.1400	14.85	1.461	0.3690	58	-0.11	009.09	0
009	Fiber, Neutral Detergent (%)	0037	14.79	0.6500	14.85	1.461	0.3690	58	-0.05	009.09	0
009	Fiber, Neutral Detergent (%)	0675	14.79	0.3000	14.85	1.461	0.3690	58	-0.04	009.09	0
009	Fiber, Neutral Detergent (%)	0278	14.80	0.8000	14.85	1.461	0.3690	58	-0.04	009.09	0
009	Fiber, Neutral Detergent (%)	0870	15.04	0.1347	14.85	1.461	0.3690	58	0.13	009.09	0
009	Fiber, Neutral Detergent (%)	0066	15.04	0.4690	14.85	1.461	0.3690	58	0.13	009.09	0
009	Fiber, Neutral Detergent (%)	0675	15.19	0.2700	14.85	1.461	0.3690	58	0.23	009.07	0
009	Fiber, Neutral Detergent (%)	0964	15.20	0.0470	14.85	1.461	0.3690	58	0.24	009.07	0
009	Fiber, Neutral Detergent (%)	0918	15.23	1.263	14.85	1.461	0.3690	58	0.26	009.09	0
009	Fiber, Neutral Detergent (%)	0510	15.25	0.1000	14.85	1.461	0.3690	58	0.27	009.09	0
009	Fiber, Neutral Detergent (%)	0083	15.61	0.1900	14.85	1.461	0.3690	58	0.51	009.09	0
009	Fiber, Neutral Detergent (%)	2129	15.63	0.1800	14.85	1.461	0.3690	58	0.53	009.09	0
009	Fiber, Neutral Detergent (%)	0032	15.78	1.010	14.85	1.461	0.3690	58	0.63	009.09	0
009	Fiber, Neutral Detergent (%)	0693	15.94	0.4420	14.85	1.461	0.3690	58	0.74	009.09	0
009	Fiber, Neutral Detergent (%)	2192	15.95	1.310	14.85	1.461	0.3690	58	0.75	009.07	0
009	Fiber, Neutral Detergent (%)	0590	15.98	0.7900	14.85	1.461	0.3690	58	0.77	009.09	0
009	Fiber, Neutral Detergent (%)	0226	16.05	0.1000	14.85	1.461	0.3690	58	0.82	009.07	0
009	Fiber, Neutral Detergent (%)	0610	16.05	0.5000	14.85	1.461	0.3690	58	0.82	009.07	0
009	Fiber, Neutral Detergent (%)	0309	16.09	0.0018	14.85	1.461	0.3690	58	0.85	009.07	0
009	Fiber, Neutral Detergent (%)	0676	16.18	0.1400	14.85	1.461	0.3690	58	0.91	009.07	0
009	Fiber, Neutral Detergent (%)	0848	16.42	0.0000	14.85	1.461	0.3690	58	1.07	009.09	0
009	Fiber, Neutral Detergent (%)	0726	16.46	0.5000	14.85	1.461	0.3690	58	1.10	009.09	0
009	Fiber, Neutral Detergent (%)	0265	16.50	0.4000	14.85	1.461	0.3690	58	1.13	009.09	0
009	Fiber, Neutral Detergent (%)	0536	16.58	0.4300	14.85	1.461	0.3690	58	1.18	009.09	0
009	Fiber, Neutral Detergent (%)	2246	16.66	0.0100	14.85	1.461	0.3690	58	1.23	009.07	0
009	Fiber, Neutral Detergent (%)	0190	16.83	0.0500	14.85	1.461	0.3690	58	1.35	009.09	0
009	Fiber, Neutral Detergent (%)	0889	17.58	0.0700	14.85	1.461	0.3690	58	1.86	009.09	0
009	Fiber, Neutral Detergent (%)	0968	17.85	1.656	14.85	1.461	0.3690	58	2.05	009.09	0
009	Fiber, Neutral Detergent (%)	0045	18.10	1.400	14.85	1.461	0.3690	58	2.22	009.09	0
009	Fiber, Neutral Detergent (%)	2246	19.74	0.4800	14.85	1.461	0.3690	58	3.34	009.04	0
009	Fiber, Neutral Detergent (%)	0656	32.92	0.8800	14.85	1.461	0.3690	58	12.36	009.09	2
010	Moisture (%)	2129	7.985	0.2700	9.033	0.5868	0.1384	21	-1.79	010.99	0
010	Moisture (%)	0720	8.075	0.0300	9.033	0.5868	0.1384	21	-1.63	010.11	0
010	Moisture (%)	0948	8.375	0.0700	9.033	0.5868	0.1384	21	-1.12	010.99	0
010	Moisture (%)	0692	8.470	0.1800	9.033	0.5868	0.1384	21	-0.96	010.99	0
010	Moisture (%)	0939	8.675	0.0500	9.033	0.5868	0.1384	21	-0.61	010.99	0
010	Moisture (%)	2251	8.705	0.0900	9.033	0.5868	0.1384	21	-0.56	010.99	0
010	Moisture (%)	0884	8.850	0.3000	9.033	0.5868	0.1384	21	-0.31	010.99	0
010	Moisture (%)	2089	8.870	0.1200	9.033	0.5868	0.1384	21	-0.28	010.99	0
010	Moisture (%)	2043	8.920	0.0800	9.033	0.5868	0.1384	21	-0.19	010.99	0
010	Moisture (%)	0726	8.925	0.0100	9.033	0.5868	0.1384	21	-0.18	010.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
010	Moisture (%)	2150	9.085	0.0700	9.033	0.5868	0.1384	21	0.09	010.99	0
010	Moisture (%)	0652	9.100	0.0000	9.033	0.5868	0.1384	21	0.11	010.99	0
010	Moisture (%)	0164	9.135	0.5700	9.033	0.5868	0.1384	21	0.17	010.03	0
010	Moisture (%)	2178	9.180	0.2000	9.033	0.5868	0.1384	21	0.25	010.99	0
010	Moisture (%)	0843	9.200	0.2400	9.033	0.5868	0.1384	21	0.29	010.03	0
010	Moisture (%)	0190	9.245	0.0500	9.033	0.5868	0.1384	21	0.36	010.99	0
010	Moisture (%)	0889	9.448	0.0097	9.033	0.5868	0.1384	21	0.71	010.11	0
010	Moisture (%)	2076	9.520	0.0800	9.033	0.5868	0.1384	21	0.83	010.99	0
010	Moisture (%)	2248	9.850	0.0000	9.033	0.5868	0.1384	21	1.39	010.99	0
010	Moisture (%)	2128	10.11	0.4067	9.033	0.5868	0.1384	21	1.84	010.11	0
010	Moisture (%)	0852	10.96	0.0800	9.033	0.5868	0.1384	21	3.28	010.11	0
011	Loss on Drying (%)	0880	7.480	0.0400	9.429	0.3217	0.0845	64	-6.06	011.01	0
011	Loss on Drying (%)	2199	8.300	0.2000	9.429	0.3217	0.0845	64	-3.51	011.01	0
011	Loss on Drying (%)	0685	8.655	0.1500	9.429	0.3217	0.0845	64	-2.41	011.01	0
011	Loss on Drying (%)	0874	8.720	0.0600	9.429	0.3217	0.0845	64	-2.20	011.01	0
011	Loss on Drying (%)	0650	8.815	0.0900	9.429	0.3217	0.0845	64	-1.91	011.01	0
011	Loss on Drying (%)	0852	8.825	0.0100	9.429	0.3217	0.0845	64	-1.88	011.99	0
011	Loss on Drying (%)	0175	8.900	0.0000	9.429	0.3217	0.0845	64	-1.64	011.01	0
011	Loss on Drying (%)	0539	9.015	0.1700	9.429	0.3217	0.0845	64	-1.29	011.01	0
011	Loss on Drying (%)	0298	9.080	0.0200	9.429	0.3217	0.0845	64	-1.08	011.01	0
011	Loss on Drying (%)	0953	9.085	0.1500	9.429	0.3217	0.0845	64	-1.07	011.01	0
011	Loss on Drying (%)	0881	9.098	0.0800	9.429	0.3217	0.0845	64	-1.03	011.01	0
011	Loss on Drying (%)	0859	9.112	0.0080	9.429	0.3217	0.0845	64	-0.99	011.01	0
011	Loss on Drying (%)	2188	9.180	0.1200	9.429	0.3217	0.0845	64	-0.77	011.01	0
011	Loss on Drying (%)	0425	9.220	0.0400	9.429	0.3217	0.0845	64	-0.65	011.01	0
011	Loss on Drying (%)	2193	9.230	0.2000	9.429	0.3217	0.0845	64	-0.62	011.01	0
011	Loss on Drying (%)	2006	9.235	0.2700	9.429	0.3217	0.0845	64	-0.60	011.01	0
011	Loss on Drying (%)	0574	9.240	0.0000	9.429	0.3217	0.0845	64	-0.59	011.01	0
011	Loss on Drying (%)	2073	9.290	0.0000	9.429	0.3217	0.0845	64	-0.43	011.01	0
011	Loss on Drying (%)	2268	9.305	0.0700	9.429	0.3217	0.0845	64	-0.39	011.01	0
011	Loss on Drying (%)	2168	9.315	0.1100	9.429	0.3217	0.0845	64	-0.35	011.01	0
011	Loss on Drying (%)	0942	9.315	0.0900	9.429	0.3217	0.0845	64	-0.35	011.02	0
011	Loss on Drying (%)	0598	9.325	0.2100	9.429	0.3217	0.0845	64	-0.32	011.01	0
011	Loss on Drying (%)	0062	9.333	0.0110	9.429	0.3217	0.0845	64	-0.30	011.01	0
011	Loss on Drying (%)	0226	9.335	0.1100	9.429	0.3217	0.0845	64	-0.29	011.01	0
011	Loss on Drying (%)	0520	9.350	0.1000	9.429	0.3217	0.0845	64	-0.25	011.01	0
011	Loss on Drying (%)	2198	9.355	0.0100	9.429	0.3217	0.0845	64	-0.23	011.01	0
011	Loss on Drying (%)	0682	9.360	0.0000	9.429	0.3217	0.0845	64	-0.21	011.01	0
011	Loss on Drying (%)	2033	9.375	0.0100	9.429	0.3217	0.0845	64	-0.17	011.01	0
011	Loss on Drying (%)	0968	9.378	0.0220	9.429	0.3217	0.0845	64	-0.16	011.01	0
011	Loss on Drying (%)	2181	9.380	0.1200	9.429	0.3217	0.0845	64	-0.15	011.01	0
011	Loss on Drying (%)	0511	9.385	0.1100	9.429	0.3217	0.0845	64	-0.14	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
011	Loss on Drying (%)	0857	9.400	0.0000	9.429	0.3217	0.0845	64	-0.09	011.99	0
011	Loss on Drying (%)	0108	9.420	0.2400	9.429	0.3217	0.0845	64	-0.03	011.01	0
011	Loss on Drying (%)	0164	9.420	0.1400	9.429	0.3217	0.0845	64	-0.03	011.01	0
011	Loss on Drying (%)	0870	9.444	0.1625	9.429	0.3217	0.0845	64	0.05	011.01	0
011	Loss on Drying (%)	0229	9.487	0.0040	9.429	0.3217	0.0845	64	0.18	011.01	0
011	Loss on Drying (%)	0026	9.505	0.0300	9.429	0.3217	0.0845	64	0.24	011.01	0
011	Loss on Drying (%)	0675	9.505	0.0100	9.429	0.3217	0.0845	64	0.24	011.01	0
011	Loss on Drying (%)	0354	9.520	0.0000	9.429	0.3217	0.0845	64	0.28	011.01	0
011	Loss on Drying (%)	2022	9.545	0.0500	9.429	0.3217	0.0845	64	0.36	011.01	0
011	Loss on Drying (%)	0265	9.550	0.1000	9.429	0.3217	0.0845	64	0.38	011.01	0
011	Loss on Drying (%)	0541	9.565	0.1100	9.429	0.3217	0.0845	64	0.42	011.01	0
011	Loss on Drying (%)	0100	9.565	0.1900	9.429	0.3217	0.0845	64	0.42	011.01	0
011	Loss on Drying (%)	0171	9.575	0.0100	9.429	0.3217	0.0845	64	0.45	011.01	0
011	Loss on Drying (%)	0144	9.595	0.0900	9.429	0.3217	0.0845	64	0.52	011.01	0
011	Loss on Drying (%)	0294	9.600	0.2000	9.429	0.3217	0.0845	64	0.53	011.01	0
011	Loss on Drying (%)	2196	9.600	0.0000	9.429	0.3217	0.0845	64	0.53	011.01	0
011	Loss on Drying (%)	0047	9.615	0.0700	9.429	0.3217	0.0845	64	0.58	011.01	0
011	Loss on Drying (%)	0098	9.630	0.1200	9.429	0.3217	0.0845	64	0.62	011.01	0
011	Loss on Drying (%)	0242	9.630	0.2800	9.429	0.3217	0.0845	64	0.62	011.01	0
011	Loss on Drying (%)	0536	9.680	0.0200	9.429	0.3217	0.0845	64	0.78	011.01	0
011	Loss on Drying (%)	0123	9.695	0.0100	9.429	0.3217	0.0845	64	0.83	011.01	0
011	Loss on Drying (%)	0510	9.700	0.2000	9.429	0.3217	0.0845	64	0.84	011.01	0
011	Loss on Drying (%)	0309	9.715	0.1100	9.429	0.3217	0.0845	64	0.89	011.01	0
011	Loss on Drying (%)	0263	9.726	0.0370	9.429	0.3217	0.0845	64	0.92	011.01	0
011	Loss on Drying (%)	0407	9.732	0.0766	9.429	0.3217	0.0845	64	0.94	011.01	0
011	Loss on Drying (%)	2246	9.830	0.0200	9.429	0.3217	0.0845	64	1.25	011.01	0
011	Loss on Drying (%)	0618	9.860	0.0200	9.429	0.3217	0.0845	64	1.34	011.01	0
011	Loss on Drying (%)	0955	9.870	0.1000	9.429	0.3217	0.0845	64	1.37	011.01	0
011	Loss on Drying (%)	0589	9.910	0.0800	9.429	0.3217	0.0845	64	1.50	011.01	0
011	Loss on Drying (%)	0202	9.996	0.1200	9.429	0.3217	0.0845	64	1.76	011.01	0
011	Loss on Drying (%)	0066	10.01	0.0900	9.429	0.3217	0.0845	64	1.79	011.01	0
011	Loss on Drying (%)	2184	10.09	0.0000	9.429	0.3217	0.0845	64	2.05	011.01	0
011	Loss on Drying (%)	0004	10.28	0.1400	9.429	0.3217	0.0845	64	2.65	011.01	0
011	Loss on Drying (%)	0553	8.765	0.9500	9.429	0.3217	0.0845	64	-2.06	011.01	1
011	Loss on Drying (%)	0643	9.050	0.7000	9.429	0.3217	0.0845	64	-1.18	011.01	1
011	Loss on Drying (%)	0843	9.495	0.8900	9.429	0.3217	0.0845	64	0.21	011.01	1
011	Loss on Drying (%)	0358	90.26	0.2000	9.429	0.3217	0.0845	64	251.23	011.01	2
012	Starch (%)	0164	24.20	0.8000	28.06	1.631	0.3630	44	-2.37	012.01	0
012	Starch (%)	0510	25.10	0.0000	28.06	1.631	0.3630	44	-1.81	012.04	0
012	Starch (%)	0407	25.58	0.1593	28.06	1.631	0.3630	44	-1.52	012.03	0
012	Starch (%)	0870	25.60	0.4446	28.06	1.631	0.3630	44	-1.51	012.01	0
012	Starch (%)	0676	25.75	0.2800	28.06	1.631	0.3630	44	-1.42	012.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
012	Starch (%)	0354	25.96	0.2400	28.06	1.631	0.3630	44	-1.29	012.00	0
012	Starch (%)	0708	26.25	0.6800	28.06	1.631	0.3630	44	-1.11	012.01	0
012	Starch (%)	0148	26.45	0.6800	28.06	1.631	0.3630	44	-0.99	012.01	0
012	Starch (%)	0353	26.47	0.1300	28.06	1.631	0.3630	44	-0.98	012.04	0
012	Starch (%)	0278	26.85	0.1000	28.06	1.631	0.3630	44	-0.74	012.04	0
012	Starch (%)	0852	26.92	0.1500	28.06	1.631	0.3630	44	-0.70	012.11	0
012	Starch (%)	0885	26.95	0.3943	28.06	1.631	0.3630	44	-0.68	012.01	0
012	Starch (%)	2242	27.07	0.4700	28.06	1.631	0.3630	44	-0.61	012.01	0
012	Starch (%)	2188	27.39	0.2500	28.06	1.631	0.3630	44	-0.41	012.00	0
012	Starch (%)	0045	27.40	0.4000	28.06	1.631	0.3630	44	-0.40	012.01	0
012	Starch (%)	0689	27.60	0.8000	28.06	1.631	0.3630	44	-0.28	012.00	0
012	Starch (%)	0720	27.62	0.2200	28.06	1.631	0.3630	44	-0.27	012.11	0
012	Starch (%)	0098	27.70	1.200	28.06	1.631	0.3630	44	-0.22	012.03	0
012	Starch (%)	0693	27.72	0.2310	28.06	1.631	0.3630	44	-0.21	012.01	0
012	Starch (%)	0656	27.86	0.0000	28.06	1.631	0.3630	44	-0.12	012.03	0
012	Starch (%)	0164	28.20	0.8000	28.06	1.631	0.3630	44	0.09	012.00	0
012	Starch (%)	2192	28.30	0.3300	28.06	1.631	0.3630	44	0.15	012.00	0
012	Starch (%)	0948	28.30	0.5500	28.06	1.631	0.3630	44	0.15	012.04	0
012	Starch (%)	0939	28.42	0.9800	28.06	1.631	0.3630	44	0.22	012.00	0
012	Starch (%)	0610	28.50	0.4000	28.06	1.631	0.3630	44	0.27	012.00	0
012	Starch (%)	2196	28.68	0.0000	28.06	1.631	0.3630	44	0.38	012.00	0
012	Starch (%)	2181	28.81	0.2100	28.06	1.631	0.3630	44	0.46	012.00	0
012	Starch (%)	0297	28.83	1.750	28.06	1.631	0.3630	44	0.47	012.03	0
012	Starch (%)	0918	28.89	0.3660	28.06	1.631	0.3630	44	0.51	012.01	0
012	Starch (%)	0619	28.90	0.2000	28.06	1.631	0.3630	44	0.52	012.00	0
012	Starch (%)	0910	28.90	0.6000	28.06	1.631	0.3630	44	0.52	012.00	0
012	Starch (%)	2193	28.91	0.1500	28.06	1.631	0.3630	44	0.52	012.00	0
012	Starch (%)	2006	29.04	0.0200	28.06	1.631	0.3630	44	0.60	012.00	0
012	Starch (%)	0889	29.05	0.5047	28.06	1.631	0.3630	44	0.61	012.03	0
012	Starch (%)	0878	29.20	0.0000	28.06	1.631	0.3630	44	0.70	012.00	0
012	Starch (%)	0941	29.48	0.1800	28.06	1.631	0.3630	44	0.87	012.00	0
012	Starch (%)	2009	29.52	0.0120	28.06	1.631	0.3630	44	0.90	012.00	0
012	Starch (%)	0889	29.76	0.0262	28.06	1.631	0.3630	44	1.05	012.11	0
012	Starch (%)	0297	29.79	0.0900	28.06	1.631	0.3630	44	1.06	012.11	0
012	Starch (%)	0683	29.84	0.1100	28.06	1.631	0.3630	44	1.09	012.00	0
012	Starch (%)	0066	29.87	0.4600	28.06	1.631	0.3630	44	1.11	012.03	0
012	Starch (%)	0265	30.00	0.4000	28.06	1.631	0.3630	44	1.19	012.01	0
012	Starch (%)	2089	31.55	0.0800	28.06	1.631	0.3630	44	2.14	012.01	0
012	Starch (%)	2128	32.45	0.1260	28.06	1.631	0.3630	44	2.69	012.11	0
012	Starch (%)	0956	26.00	1.800	28.06	1.631	0.3630	44	-1.26	012.01	1
013	Fat, Acid Pretreat (%)	0618	3.375	0.0134	4.655	0.5623	0.1179	49	-2.28	013.08	0
013	Fat, Acid Pretreat (%)	0036	3.668	0.1321	4.655	0.5623	0.1179	49	-1.75	013.10	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
013	Fat, Acid Pretreat (%)	0581	3.745	0.0300	4.655	0.5623	0.1179	49	-1.62	013.13	0
013	Fat, Acid Pretreat (%)	0874	3.780	0.1600	4.655	0.5623	0.1179	49	-1.56	013.02	0
013	Fat, Acid Pretreat (%)	0083	3.870	0.0200	4.655	0.5623	0.1179	49	-1.40	013.00	0
013	Fat, Acid Pretreat (%)	0242	4.055	0.1700	4.655	0.5623	0.1179	49	-1.07	013.00	0
013	Fat, Acid Pretreat (%)	2073	4.085	0.0100	4.655	0.5623	0.1179	49	-1.01	013.00	0
013	Fat, Acid Pretreat (%)	0353	4.105	0.0100	4.655	0.5623	0.1179	49	-0.98	013.10	0
013	Fat, Acid Pretreat (%)	0297	4.120	0.0200	4.655	0.5623	0.1179	49	-0.95	013.00	0
013	Fat, Acid Pretreat (%)	0504	4.150	0.0200	4.655	0.5623	0.1179	49	-0.90	013.00	0
013	Fat, Acid Pretreat (%)	0425	4.190	0.1200	4.655	0.5623	0.1179	49	-0.83	013.02	0
013	Fat, Acid Pretreat (%)	0650	4.225	0.0900	4.655	0.5623	0.1179	49	-0.76	013.02	0
013	Fat, Acid Pretreat (%)	0853	4.245	0.0700	4.655	0.5623	0.1179	49	-0.73	013.02	0
013	Fat, Acid Pretreat (%)	0720	4.290	0.0800	4.655	0.5623	0.1179	49	-0.65	013.10	0
013	Fat, Acid Pretreat (%)	0309	4.355	0.1348	4.655	0.5623	0.1179	49	-0.53	013.00	0
013	Fat, Acid Pretreat (%)	0689	4.400	0.2000	4.655	0.5623	0.1179	49	-0.45	013.00	0
013	Fat, Acid Pretreat (%)	0610	4.480	0.1000	4.655	0.5623	0.1179	49	-0.31	013.10	0
013	Fat, Acid Pretreat (%)	0407	4.495	0.1300	4.655	0.5623	0.1179	49	-0.28	013.02	0
013	Fat, Acid Pretreat (%)	0948	4.505	0.1500	4.655	0.5623	0.1179	49	-0.27	013.00	0
013	Fat, Acid Pretreat (%)	0229	4.510	0.3600	4.655	0.5623	0.1179	49	-0.26	013.00	0
013	Fat, Acid Pretreat (%)	0910	4.510	0.1200	4.655	0.5623	0.1179	49	-0.26	013.00	0
013	Fat, Acid Pretreat (%)	2196	4.530	0.0000	4.655	0.5623	0.1179	49	-0.22	013.10	0
013	Fat, Acid Pretreat (%)	0652	4.550	0.1000	4.655	0.5623	0.1179	49	-0.19	013.00	0
013	Fat, Acid Pretreat (%)	0074	4.560	0.3600	4.655	0.5623	0.1179	49	-0.17	013.02	0
013	Fat, Acid Pretreat (%)	0541	4.590	0.4200	4.655	0.5623	0.1179	49	-0.11	013.00	0
013	Fat, Acid Pretreat (%)	0098	4.655	0.0700	4.655	0.5623	0.1179	49	0.00	013.02	0
013	Fat, Acid Pretreat (%)	0884	4.700	0.0000	4.655	0.5623	0.1179	49	0.08	013.00	0
013	Fat, Acid Pretreat (%)	0164	4.715	0.1100	4.655	0.5623	0.1179	49	0.11	013.02	0
013	Fat, Acid Pretreat (%)	0870	4.834	0.0024	4.655	0.5623	0.1179	49	0.32	013.02	0
013	Fat, Acid Pretreat (%)	0354	4.860	0.0600	4.655	0.5623	0.1179	49	0.37	013.02	0
013	Fat, Acid Pretreat (%)	2264	4.887	0.0426	4.655	0.5623	0.1179	49	0.41	013.02	0
013	Fat, Acid Pretreat (%)	2033	4.905	0.0500	4.655	0.5623	0.1179	49	0.45	013.02	0
013	Fat, Acid Pretreat (%)	2076	4.987	0.0035	4.655	0.5623	0.1179	49	0.59	013.00	0
013	Fat, Acid Pretreat (%)	0407	5.013	0.6379	4.655	0.5623	0.1179	49	0.64	013.13	0
013	Fat, Acid Pretreat (%)	0643	5.045	0.2700	4.655	0.5623	0.1179	49	0.69	013.02	0
013	Fat, Acid Pretreat (%)	2128	5.064	0.0333	4.655	0.5623	0.1179	49	0.73	013.12	0
013	Fat, Acid Pretreat (%)	0618	5.087	0.0484	4.655	0.5623	0.1179	49	0.77	013.00	0
013	Fat, Acid Pretreat (%)	2150	5.100	0.0800	4.655	0.5623	0.1179	49	0.79	013.00	0
013	Fat, Acid Pretreat (%)	0553	5.105	0.1500	4.655	0.5623	0.1179	49	0.80	013.02	0
013	Fat, Acid Pretreat (%)	0148	5.110	0.0000	4.655	0.5623	0.1179	49	0.81	013.02	0
013	Fat, Acid Pretreat (%)	0682	5.120	0.0000	4.655	0.5623	0.1179	49	0.83	013.02	0
013	Fat, Acid Pretreat (%)	0675	5.160	0.0200	4.655	0.5623	0.1179	49	0.90	013.02	0
013	Fat, Acid Pretreat (%)	2259	5.170	0.1460	4.655	0.5623	0.1179	49	0.92	013.02	0
013	Fat, Acid Pretreat (%)	0100	5.285	0.0700	4.655	0.5623	0.1179	49	1.12	013.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
013	Fat, Acid Pretreat (%)	0202	5.318	0.0850	4.655	0.5623	0.1179	49	1.18	013.00	0
013	Fat, Acid Pretreat (%)	0171	5.435	0.4700	4.655	0.5623	0.1179	49	1.39	013.02	0
013	Fat, Acid Pretreat (%)	0265	5.455	0.0300	4.655	0.5623	0.1179	49	1.42	013.13	0
013	Fat, Acid Pretreat (%)	0676	5.585	0.2500	4.655	0.5623	0.1179	49	1.65	013.02	0
013	Fat, Acid Pretreat (%)	0955	5.715	0.1300	4.655	0.5623	0.1179	49	1.89	013.13	0
013	Fat, Acid Pretreat (%)	0843	5.475	0.9100	4.655	0.5623	0.1179	49	1.46	013.13	1
015	Aluminum (mg / kg (ppm))	0918	106.0	5.046	137.7	20.53	6.914	16	-1.55	015.43	0
015	Aluminum (mg / kg (ppm))	0510	118.0	0.0000	137.7	20.53	6.914	16	-0.96	015.43	0
015	Aluminum (mg / kg (ppm))	0560	122.7	10.10	137.7	20.53	6.914	16	-0.74	015.42	0
015	Aluminum (mg / kg (ppm))	0169	123.5	3.000	137.7	20.53	6.914	16	-0.69	015.43	0
015	Aluminum (mg / kg (ppm))	0049	124.5	7.450	137.7	20.53	6.914	16	-0.65	015.41	0
015	Aluminum (mg / kg (ppm))	0407	124.8	2.275	137.7	20.53	6.914	16	-0.63	015.41	0
015	Aluminum (mg / kg (ppm))	0425	127.2	2.260	137.7	20.53	6.914	16	-0.52	015.43	0
015	Aluminum (mg / kg (ppm))	0345	130.5	3.000	137.7	20.53	6.914	16	-0.35	015.43	0
015	Aluminum (mg / kg (ppm))	0171	139.9	2.300	137.7	20.53	6.914	16	0.10	015.41	0
015	Aluminum (mg / kg (ppm))	0032	140.5	11.00	137.7	20.53	6.914	16	0.13	015.43	0
015	Aluminum (mg / kg (ppm))	0297	146.5	11.00	137.7	20.53	6.914	16	0.43	015.43	0
015	Aluminum (mg / kg (ppm))	0353	154.0	8.000	137.7	20.53	6.914	16	0.79	015.43	0
015	Aluminum (mg / kg (ppm))	0164	156.5	17.00	137.7	20.53	6.914	16	0.91	015.41	0
015	Aluminum (mg / kg (ppm))	0407	159.9	8.196	137.7	20.53	6.914	16	1.08	015.43	0
015	Aluminum (mg / kg (ppm))	0553	160.0	16.00	137.7	20.53	6.914	16	1.08	015.53	0
015	Aluminum (mg / kg (ppm))	0520	171.0	4.000	137.7	20.53	6.914	16	1.62	015.41	0
017	Boron (mg / kg (ppm))	0297	2.000	0.0000	9.869	1.560	0.3774	20	-5.04	017.43	0
017	Boron (mg / kg (ppm))	0229	6.715	0.3900	9.869	1.560	0.3774	20	-2.02	017.41	0
017	Boron (mg / kg (ppm))	0032	7.510	0.3600	9.869	1.560	0.3774	20	-1.51	017.43	0
017	Boron (mg / kg (ppm))	0265	8.065	0.1700	9.869	1.560	0.3774	20	-1.16	017.42	0
017	Boron (mg / kg (ppm))	0345	9.200	0.2000	9.869	1.560	0.3774	20	-0.43	017.43	0
017	Boron (mg / kg (ppm))	0553	9.435	0.4900	9.869	1.560	0.3774	20	-0.28	017.53	0
017	Boron (mg / kg (ppm))	0693	9.442	0.1860	9.869	1.560	0.3774	20	-0.27	017.42	0
017	Boron (mg / kg (ppm))	0358	9.915	1.610	9.869	1.560	0.3774	20	0.03	017.41	0
017	Boron (mg / kg (ppm))	0510	10.00	0.0000	9.869	1.560	0.3774	20	0.08	017.43	0
017	Boron (mg / kg (ppm))	0918	10.08	0.0100	9.869	1.560	0.3774	20	0.13	017.43	0
017	Boron (mg / kg (ppm))	0226	10.16	0.0300	9.869	1.560	0.3774	20	0.18	017.41	0
017	Boron (mg / kg (ppm))	0045	10.20	0.2000	9.869	1.560	0.3774	20	0.21	017.42	0
017	Boron (mg / kg (ppm))	0560	10.25	0.6000	9.869	1.560	0.3774	20	0.24	017.42	0
017	Boron (mg / kg (ppm))	0407	10.40	0.6769	9.869	1.560	0.3774	20	0.34	017.41	0
017	Boron (mg / kg (ppm))	0353	10.50	1.000	9.869	1.560	0.3774	20	0.40	017.43	0
017	Boron (mg / kg (ppm))	0407	10.61	1.325	9.869	1.560	0.3774	20	0.47	017.43	0
017	Boron (mg / kg (ppm))	0083	10.95	0.1000	9.869	1.560	0.3774	20	0.69	017.43	0
017	Boron (mg / kg (ppm))	2053	11.18	0.0300	9.869	1.560	0.3774	20	0.84	017.42	0
017	Boron (mg / kg (ppm))	0294	12.94	0.0800	9.869	1.560	0.3774	20	1.97	017.42	0
017	Boron (mg / kg (ppm))	0049	12.99	0.0900	9.869	1.560	0.3774	20	2.00	017.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	2097	0.1365	0.0070	4.891	0.2426	0.0928	134	-19.60	019.31	0
019	Calcium (%)	2199	3.550	0.1000	4.891	0.2426	0.0928	134	-5.53	019.08	0
019	Calcium (%)	2200	3.600	0.2000	4.891	0.2426	0.0928	134	-5.32	019.08	0
019	Calcium (%)	0889	3.677	0.0380	4.891	0.2426	0.0928	134	-5.00	019.99	0
019	Calcium (%)	0536	3.965	0.0700	4.891	0.2426	0.0928	134	-3.82	019.02	0
019	Calcium (%)	0123	4.135	0.0700	4.891	0.2426	0.0928	134	-3.11	019.41	0
019	Calcium (%)	0089	4.305	0.0100	4.891	0.2426	0.0928	134	-2.41	019.43	0
019	Calcium (%)	0298	4.440	0.0200	4.891	0.2426	0.0928	134	-1.86	019.41	0
019	Calcium (%)	0154	4.513	0.0279	4.891	0.2426	0.0928	134	-1.56	019.52	0
019	Calcium (%)	0910	4.540	0.2600	4.891	0.2426	0.0928	134	-1.45	019.41	0
019	Calcium (%)	0144	4.585	0.1300	4.891	0.2426	0.0928	134	-1.26	019.41	0
019	Calcium (%)	0265	4.595	0.2900	4.891	0.2426	0.0928	134	-1.22	019.42	0
019	Calcium (%)	2062	4.596	0.0370	4.891	0.2426	0.0928	134	-1.22	019.00	0
019	Calcium (%)	0294	4.600	0.0400	4.891	0.2426	0.0928	134	-1.20	019.42	0
019	Calcium (%)	0036	4.637	0.0519	4.891	0.2426	0.0928	134	-1.05	019.43	0
019	Calcium (%)	2089	4.640	0.2000	4.891	0.2426	0.0928	134	-1.03	019.43	0
019	Calcium (%)	0590	4.645	0.0100	4.891	0.2426	0.0928	134	-1.01	019.08	0
019	Calcium (%)	0035	4.645	0.0500	4.891	0.2426	0.0928	134	-1.01	019.42	0
019	Calcium (%)	0098	4.650	0.2600	4.891	0.2426	0.0928	134	-0.99	019.43	0
019	Calcium (%)	2192	4.650	0.2400	4.891	0.2426	0.0928	134	-0.99	019.43	0
019	Calcium (%)	0242	4.655	0.0500	4.891	0.2426	0.0928	134	-0.97	019.99	0
019	Calcium (%)	0098	4.660	0.0400	4.891	0.2426	0.0928	134	-0.95	019.41	0
019	Calcium (%)	0848	4.670	0.0000	4.891	0.2426	0.0928	134	-0.91	019.41	0
019	Calcium (%)	0037	4.671	0.0930	4.891	0.2426	0.0928	134	-0.91	019.42	0
019	Calcium (%)	0148	4.680	0.1400	4.891	0.2426	0.0928	134	-0.87	019.41	0
019	Calcium (%)	0885	4.697	0.0394	4.891	0.2426	0.0928	134	-0.80	019.08	0
019	Calcium (%)	0202	4.698	0.0830	4.891	0.2426	0.0928	134	-0.80	019.43	0
019	Calcium (%)	0004	4.700	0.3800	4.891	0.2426	0.0928	134	-0.79	019.41	0
019	Calcium (%)	0629	4.700	0.0200	4.891	0.2426	0.0928	134	-0.79	019.08	0
019	Calcium (%)	2022	4.700	0.0200	4.891	0.2426	0.0928	134	-0.79	019.31	0
019	Calcium (%)	0019	4.705	0.0300	4.891	0.2426	0.0928	134	-0.76	019.41	0
019	Calcium (%)	0536	4.716	0.0230	4.891	0.2426	0.0928	134	-0.72	019.31	0
019	Calcium (%)	0675	4.720	0.0200	4.891	0.2426	0.0928	134	-0.70	019.43	0
019	Calcium (%)	0169	4.725	0.0500	4.891	0.2426	0.0928	134	-0.68	019.32	0
019	Calcium (%)	2188	4.730	0.0000	4.891	0.2426	0.0928	134	-0.66	019.08	0
019	Calcium (%)	0171	4.735	0.0500	4.891	0.2426	0.0928	134	-0.64	019.41	0
019	Calcium (%)	0425	4.745	0.0700	4.891	0.2426	0.0928	134	-0.60	019.43	0
019	Calcium (%)	0345	4.750	0.0110	4.891	0.2426	0.0928	134	-0.58	019.43	0
019	Calcium (%)	2018	4.760	0.0200	4.891	0.2426	0.0928	134	-0.54	019.31	0
019	Calcium (%)	0358	4.770	0.1800	4.891	0.2426	0.0928	134	-0.50	019.41	0
019	Calcium (%)	0353	4.770	0.0800	4.891	0.2426	0.0928	134	-0.50	019.43	0
019	Calcium (%)	0297	4.775	0.0500	4.891	0.2426	0.0928	134	-0.48	019.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	0142	4.780	0.1035	4.891	0.2426	0.0928	134	-0.46	019.31	0
019	Calcium (%)	0357	4.783	0.0132	4.891	0.2426	0.0928	134	-0.45	019.42	0
019	Calcium (%)	0175	4.785	0.1300	4.891	0.2426	0.0928	134	-0.44	019.31	0
019	Calcium (%)	0010	4.795	0.0300	4.891	0.2426	0.0928	134	-0.39	019.43	0
019	Calcium (%)	0560	4.796	0.2230	4.891	0.2426	0.0928	134	-0.39	019.42	0
019	Calcium (%)	0968	4.796	0.0100	4.891	0.2426	0.0928	134	-0.39	019.43	0
019	Calcium (%)	0164	4.800	0.0800	4.891	0.2426	0.0928	134	-0.37	019.41	0
019	Calcium (%)	0278	4.800	0.1200	4.891	0.2426	0.0928	134	-0.37	019.42	0
019	Calcium (%)	0598	4.810	0.0470	4.891	0.2426	0.0928	134	-0.33	019.41	0
019	Calcium (%)	0510	4.810	0.1000	4.891	0.2426	0.0928	134	-0.33	019.43	0
019	Calcium (%)	2109	4.816	0.0870	4.891	0.2426	0.0928	134	-0.31	019.00	0
019	Calcium (%)	2259	4.820	0.1080	4.891	0.2426	0.0928	134	-0.29	019.41	0
019	Calcium (%)	0083	4.830	0.2200	4.891	0.2426	0.0928	134	-0.25	019.43	0
019	Calcium (%)	0003	4.835	0.0500	4.891	0.2426	0.0928	134	-0.23	019.41	0
019	Calcium (%)	2193	4.840	0.0000	4.891	0.2426	0.0928	134	-0.21	019.08	0
019	Calcium (%)	0108	4.840	0.0600	4.891	0.2426	0.0928	134	-0.21	019.31	0
019	Calcium (%)	0038	4.849	0.0920	4.891	0.2426	0.0928	134	-0.17	019.31	0
019	Calcium (%)	0689	4.850	0.0000	4.891	0.2426	0.0928	134	-0.17	019.08	0
019	Calcium (%)	2129	4.850	0.4000	4.891	0.2426	0.0928	134	-0.17	019.42	0
019	Calcium (%)	0849	4.852	0.0240	4.891	0.2426	0.0928	134	-0.16	019.08	0
019	Calcium (%)	0683	4.855	0.1300	4.891	0.2426	0.0928	134	-0.15	019.00	0
019	Calcium (%)	0066	4.856	0.0900	4.891	0.2426	0.0928	134	-0.14	019.41	0
019	Calcium (%)	0897	4.875	0.0300	4.891	0.2426	0.0928	134	-0.06	019.00	0
019	Calcium (%)	0918	4.876	0.0860	4.891	0.2426	0.0928	134	-0.06	019.43	0
019	Calcium (%)	2198	4.885	0.0500	4.891	0.2426	0.0928	134	-0.02	019.31	0
019	Calcium (%)	0354	4.885	0.0100	4.891	0.2426	0.0928	134	-0.02	019.41	0
019	Calcium (%)	2259	4.885	0.1080	4.891	0.2426	0.0928	134	-0.02	019.42	0
019	Calcium (%)	0878	4.900	0.0400	4.891	0.2426	0.0928	134	0.04	019.00	0
019	Calcium (%)	0687	4.900	0.2000	4.891	0.2426	0.0928	134	0.04	019.31	0
019	Calcium (%)	0904	4.900	0.0600	4.891	0.2426	0.0928	134	0.04	019.31	0
019	Calcium (%)	2196	4.900	0.0000	4.891	0.2426	0.0928	134	0.04	019.31	0
019	Calcium (%)	0074	4.900	0.0200	4.891	0.2426	0.0928	134	0.04	019.41	0
019	Calcium (%)	0878	4.900	0.1000	4.891	0.2426	0.0928	134	0.04	019.41	0
019	Calcium (%)	0407	4.904	0.0555	4.891	0.2426	0.0928	134	0.05	019.43	0
019	Calcium (%)	0893	4.910	0.0400	4.891	0.2426	0.0928	134	0.08	019.00	0
019	Calcium (%)	2181	4.910	0.1000	4.891	0.2426	0.0928	134	0.08	019.00	0
019	Calcium (%)	0038	4.910	0.1080	4.891	0.2426	0.0928	134	0.08	019.43	0
019	Calcium (%)	0229	4.925	0.0200	4.891	0.2426	0.0928	134	0.14	019.41	0
019	Calcium (%)	2196	4.930	0.0000	4.891	0.2426	0.0928	134	0.16	019.08	0
019	Calcium (%)	0939	4.940	0.0800	4.891	0.2426	0.0928	134	0.20	019.31	0
019	Calcium (%)	0520	4.950	0.1000	4.891	0.2426	0.0928	134	0.25	019.41	0
019	Calcium (%)	0610	4.950	0.0800	4.891	0.2426	0.0928	134	0.25	019.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	0881	4.951	0.0770	4.891	0.2426	0.0928	134	0.25	019.00	0
019	Calcium (%)	0190	4.955	0.0100	4.891	0.2426	0.0928	134	0.27	019.42	0
019	Calcium (%)	0017	4.956	0.0040	4.891	0.2426	0.0928	134	0.27	019.43	0
019	Calcium (%)	0726	4.965	0.0060	4.891	0.2426	0.0928	134	0.31	019.42	0
019	Calcium (%)	2246	4.970	0.0800	4.891	0.2426	0.0928	134	0.33	019.00	0
019	Calcium (%)	0049	4.980	0.0800	4.891	0.2426	0.0928	134	0.37	019.41	0
019	Calcium (%)	0685	4.985	0.0100	4.891	0.2426	0.0928	134	0.39	019.41	0
019	Calcium (%)	0964	4.987	0.0814	4.891	0.2426	0.0928	134	0.40	019.43	0
019	Calcium (%)	2126	4.990	0.0000	4.891	0.2426	0.0928	134	0.41	019.34	0
019	Calcium (%)	0036	4.991	0.0006	4.891	0.2426	0.0928	134	0.42	019.03	0
019	Calcium (%)	0676	4.995	0.0500	4.891	0.2426	0.0928	134	0.43	019.99	0
019	Calcium (%)	0511	5.010	0.3600	4.891	0.2426	0.0928	134	0.49	019.41	0
019	Calcium (%)	0682	5.020	0.0000	4.891	0.2426	0.0928	134	0.53	019.43	0
019	Calcium (%)	0693	5.028	0.1360	4.891	0.2426	0.0928	134	0.57	019.42	0
019	Calcium (%)	0590	5.035	0.4700	4.891	0.2426	0.0928	134	0.60	019.99	0
019	Calcium (%)	2246	5.040	0.2000	4.891	0.2426	0.0928	134	0.62	019.31	0
019	Calcium (%)	0042	5.040	0.4200	4.891	0.2426	0.0928	134	0.62	019.42	0
019	Calcium (%)	2251	5.040	0.0200	4.891	0.2426	0.0928	134	0.62	019.99	0
019	Calcium (%)	2009	5.041	0.0080	4.891	0.2426	0.0928	134	0.62	019.08	0
019	Calcium (%)	0868	5.050	0.0600	4.891	0.2426	0.0928	134	0.66	019.31	0
019	Calcium (%)	0948	5.053	0.2320	4.891	0.2426	0.0928	134	0.67	019.31	0
019	Calcium (%)	0629	5.055	0.0300	4.891	0.2426	0.0928	134	0.68	019.43	0
019	Calcium (%)	0870	5.063	0.0710	4.891	0.2426	0.0928	134	0.71	019.42	0
019	Calcium (%)	2190	5.068	0.0842	4.891	0.2426	0.0928	134	0.73	019.08	0
019	Calcium (%)	0027	5.071	0.2530	4.891	0.2426	0.0928	134	0.74	019.43	0
019	Calcium (%)	0100	5.125	0.2500	4.891	0.2426	0.0928	134	0.97	019.99	0
019	Calcium (%)	0692	5.135	0.0900	4.891	0.2426	0.0928	134	1.01	019.42	0
019	Calcium (%)	0553	5.145	0.0500	4.891	0.2426	0.0928	134	1.05	019.53	0
019	Calcium (%)	0001	5.173	0.0100	4.891	0.2426	0.0928	134	1.16	019.31	0
019	Calcium (%)	0187	5.180	0.0200	4.891	0.2426	0.0928	134	1.19	019.42	0
019	Calcium (%)	0619	5.185	0.0100	4.891	0.2426	0.0928	134	1.21	019.41	0
019	Calcium (%)	2006	5.193	0.1390	4.891	0.2426	0.0928	134	1.24	019.00	0
019	Calcium (%)	0407	5.207	0.1560	4.891	0.2426	0.0928	134	1.30	019.41	0
019	Calcium (%)	0045	5.225	0.1700	4.891	0.2426	0.0928	134	1.38	019.42	0
019	Calcium (%)	2248	5.230	0.1470	4.891	0.2426	0.0928	134	1.40	019.31	0
019	Calcium (%)	0035	5.230	0.0800	4.891	0.2426	0.0928	134	1.40	019.43	0
019	Calcium (%)	0226	5.235	0.1700	4.891	0.2426	0.0928	134	1.42	019.41	0
019	Calcium (%)	0695	5.250	0.1270	4.891	0.2426	0.0928	134	1.48	019.41	0
019	Calcium (%)	0612	5.260	0.2400	4.891	0.2426	0.0928	134	1.52	019.33	0
019	Calcium (%)	0032	5.325	0.3700	4.891	0.2426	0.0928	134	1.79	019.43	0
019	Calcium (%)	0939	5.365	0.1100	4.891	0.2426	0.0928	134	1.96	019.53	0
019	Calcium (%)	0656	5.375	0.0100	4.891	0.2426	0.0928	134	2.00	019.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	2053	5.400	0.0000	4.891	0.2426	0.0928	134	2.10	019.42	0
019	Calcium (%)	0047	5.430	0.0800	4.891	0.2426	0.0928	134	2.22	019.52	0
019	Calcium (%)	0529	5.655	0.1500	4.891	0.2426	0.0928	134	3.15	019.31	0
019	Calcium (%)	0874	5.810	0.0200	4.891	0.2426	0.0928	134	3.79	019.31	0
019	Calcium (%)	0852	5.850	0.1000	4.891	0.2426	0.0928	134	3.95	019.08	0
019	Calcium (%)	0720	5.875	0.0100	4.891	0.2426	0.0928	134	4.06	019.41	0
019	Calcium (%)	2234	6.560	0.1200	4.891	0.2426	0.0928	134	6.88	019.43	0
019	Calcium (%)	0505	7.920	0.1000	4.891	0.2426	0.0928	134	12.49	019.43	0
019	Calcium (%)	0026	3.446	0.7790	4.891	0.2426	0.0928	134	-5.96	019.42	1
019	Calcium (%)	2128	4.380	0.5850	4.891	0.2426	0.0928	134	-2.11	019.02	1
019	Calcium (%)	0366	4.890	0.5200	4.891	0.2426	0.0928	134	0.00	019.42	1
019	Calcium (%)	0011	14.59	0.0000	4.891	0.2426	0.0928	134	39.98	019.41	2
021	Cobalt (mg / kg (ppm))	0047	0.1305	0.0250	1.378	0.4298	0.0851	25	-2.90	021.52	0
021	Cobalt (mg / kg (ppm))	0910	0.6350	0.0700	1.378	0.4298	0.0851	25	-1.73	021.52	0
021	Cobalt (mg / kg (ppm))	0689	0.8500	0.1000	1.378	0.4298	0.0851	25	-1.23	021.31	0
021	Cobalt (mg / kg (ppm))	0918	0.8502	0.0584	1.378	0.4298	0.0851	25	-1.23	021.53	0
021	Cobalt (mg / kg (ppm))	0619	0.8750	0.0380	1.378	0.4298	0.0851	25	-1.17	021.41	0
021	Cobalt (mg / kg (ppm))	0345	1.085	0.0300	1.378	0.4298	0.0851	25	-0.68	021.43	0
021	Cobalt (mg / kg (ppm))	0353	1.119	0.1422	1.378	0.4298	0.0851	25	-0.60	021.43	0
021	Cobalt (mg / kg (ppm))	0693	1.137	0.3330	1.378	0.4298	0.0851	25	-0.56	021.42	0
021	Cobalt (mg / kg (ppm))	0560	1.220	0.3600	1.378	0.4298	0.0851	25	-0.37	021.42	0
021	Cobalt (mg / kg (ppm))	0964	1.244	0.0459	1.378	0.4298	0.0851	25	-0.31	021.43	0
021	Cobalt (mg / kg (ppm))	0045	1.315	0.0100	1.378	0.4298	0.0851	25	-0.15	021.42	0
021	Cobalt (mg / kg (ppm))	0032	1.425	0.3700	1.378	0.4298	0.0851	25	0.11	021.43	0
021	Cobalt (mg / kg (ppm))	0941	1.480	0.0000	1.378	0.4298	0.0851	25	0.24	021.43	0
021	Cobalt (mg / kg (ppm))	2248	1.485	0.0900	1.378	0.4298	0.0851	25	0.25	021.31	0
021	Cobalt (mg / kg (ppm))	0169	1.485	0.1500	1.378	0.4298	0.0851	25	0.25	021.43	0
021	Cobalt (mg / kg (ppm))	0148	1.580	0.0200	1.378	0.4298	0.0851	25	0.47	021.41	0
021	Cobalt (mg / kg (ppm))	0939	1.585	0.0100	1.378	0.4298	0.0851	25	0.48	021.31	0
021	Cobalt (mg / kg (ppm))	0553	1.635	0.0500	1.378	0.4298	0.0851	25	0.60	021.53	0
021	Cobalt (mg / kg (ppm))	0407	1.656	0.0033	1.378	0.4298	0.0851	25	0.65	021.41	0
021	Cobalt (mg / kg (ppm))	0164	1.700	0.0000	1.378	0.4298	0.0851	25	0.75	021.31	0
021	Cobalt (mg / kg (ppm))	0939	1.725	0.0500	1.378	0.4298	0.0851	25	0.81	021.53	0
021	Cobalt (mg / kg (ppm))	0171	1.770	0.0000	1.378	0.4298	0.0851	25	0.91	021.41	0
021	Cobalt (mg / kg (ppm))	0510	1.805	0.1100	1.378	0.4298	0.0851	25	0.99	021.43	0
021	Cobalt (mg / kg (ppm))	0407	1.934	0.0626	1.378	0.4298	0.0851	25	1.29	021.43	0
021	Cobalt (mg / kg (ppm))	0011	4.517	0.0000	1.378	0.4298	0.0851	25	7.30	021.41	0
022	Copper (mg / kg (ppm))	2246	17.66	0.1700	25.97	1.803	0.7886	75	-4.61	022.31	0
022	Copper (mg / kg (ppm))	0536	21.63	0.1500	25.97	1.803	0.7886	75	-2.41	022.31	0
022	Copper (mg / kg (ppm))	0720	23.06	0.0000	25.97	1.803	0.7886	75	-1.62	022.41	0
022	Copper (mg / kg (ppm))	0726	23.33	1.050	25.97	1.803	0.7886	75	-1.47	022.42	0
022	Copper (mg / kg (ppm))	0017	23.34	2.341	25.97	1.803	0.7886	75	-1.46	022.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
022	Copper (mg / kg (ppm))	0265	23.45	0.1000	25.97	1.803	0.7886	75	-1.40	022.42	0
022	Copper (mg / kg (ppm))	0968	23.50	1.000	25.97	1.803	0.7886	75	-1.37	022.43	0
022	Copper (mg / kg (ppm))	0596	23.71	0.2600	25.97	1.803	0.7886	75	-1.26	022.31	0
022	Copper (mg / kg (ppm))	0345	24.20	0.2000	25.97	1.803	0.7886	75	-0.98	022.43	0
022	Copper (mg / kg (ppm))	0619	24.25	1.100	25.97	1.803	0.7886	75	-0.96	022.41	0
022	Copper (mg / kg (ppm))	0407	24.35	0.0068	25.97	1.803	0.7886	75	-0.90	022.43	0
022	Copper (mg / kg (ppm))	0032	24.40	1.000	25.97	1.803	0.7886	75	-0.87	022.43	0
022	Copper (mg / kg (ppm))	0553	24.40	0.8000	25.97	1.803	0.7886	75	-0.87	022.53	0
022	Copper (mg / kg (ppm))	0169	24.45	0.1000	25.97	1.803	0.7886	75	-0.85	022.43	0
022	Copper (mg / kg (ppm))	0939	24.46	0.6500	25.97	1.803	0.7886	75	-0.84	022.53	0
022	Copper (mg / kg (ppm))	0520	24.50	1.000	25.97	1.803	0.7886	75	-0.82	022.41	0
022	Copper (mg / kg (ppm))	0366	24.50	1.000	25.97	1.803	0.7886	75	-0.82	022.42	0
022	Copper (mg / kg (ppm))	0353	24.50	1.000	25.97	1.803	0.7886	75	-0.82	022.43	0
022	Copper (mg / kg (ppm))	0242	24.50	1.000	25.97	1.803	0.7886	75	-0.82	022.99	0
022	Copper (mg / kg (ppm))	0675	24.58	1.300	25.97	1.803	0.7886	75	-0.77	022.43	0
022	Copper (mg / kg (ppm))	0037	24.70	0.2000	25.97	1.803	0.7886	75	-0.71	022.42	0
022	Copper (mg / kg (ppm))	0354	24.93	0.1200	25.97	1.803	0.7886	75	-0.58	022.41	0
022	Copper (mg / kg (ppm))	0878	25.00	2.000	25.97	1.803	0.7886	75	-0.54	022.41	0
022	Copper (mg / kg (ppm))	0910	25.00	0.0000	25.97	1.803	0.7886	75	-0.54	022.41	0
022	Copper (mg / kg (ppm))	0100	25.00	2.000	25.97	1.803	0.7886	75	-0.54	022.99	0
022	Copper (mg / kg (ppm))	0045	25.05	0.1000	25.97	1.803	0.7886	75	-0.51	022.42	0
022	Copper (mg / kg (ppm))	0407	25.14	0.4977	25.97	1.803	0.7886	75	-0.46	022.41	0
022	Copper (mg / kg (ppm))	0171	25.15	0.1000	25.97	1.803	0.7886	75	-0.46	022.41	0
022	Copper (mg / kg (ppm))	2089	25.32	0.0600	25.97	1.803	0.7886	75	-0.36	022.43	0
022	Copper (mg / kg (ppm))	0019	25.41	0.2400	25.97	1.803	0.7886	75	-0.31	022.41	0
022	Copper (mg / kg (ppm))	0939	25.42	0.0500	25.97	1.803	0.7886	75	-0.31	022.31	0
022	Copper (mg / kg (ppm))	0098	25.42	1.600	25.97	1.803	0.7886	75	-0.31	022.43	0
022	Copper (mg / kg (ppm))	2129	25.65	2.810	25.97	1.803	0.7886	75	-0.18	022.99	0
022	Copper (mg / kg (ppm))	0190	25.67	0.5400	25.97	1.803	0.7886	75	-0.17	022.42	0
022	Copper (mg / kg (ppm))	0918	25.87	0.1400	25.97	1.803	0.7886	75	-0.06	022.43	0
022	Copper (mg / kg (ppm))	0629	25.95	1.500	25.97	1.803	0.7886	75	-0.01	022.43	0
022	Copper (mg / kg (ppm))	0229	26.00	1.490	25.97	1.803	0.7886	75	0.01	022.41	0
022	Copper (mg / kg (ppm))	0510	26.00	0.0000	25.97	1.803	0.7886	75	0.01	022.43	0
022	Copper (mg / kg (ppm))	0682	26.00	0.0000	25.97	1.803	0.7886	75	0.01	022.43	0
022	Copper (mg / kg (ppm))	0629	26.05	1.100	25.97	1.803	0.7886	75	0.04	022.31	0
022	Copper (mg / kg (ppm))	0685	26.07	0.3700	25.97	1.803	0.7886	75	0.05	022.42	0
022	Copper (mg / kg (ppm))	0598	26.07	0.5800	25.97	1.803	0.7886	75	0.05	022.41	0
022	Copper (mg / kg (ppm))	0964	26.13	0.2727	25.97	1.803	0.7886	75	0.09	022.43	0
022	Copper (mg / kg (ppm))	0042	26.25	3.700	25.97	1.803	0.7886	75	0.15	022.42	0
022	Copper (mg / kg (ppm))	0098	26.44	0.0400	25.97	1.803	0.7886	75	0.26	022.41	0
022	Copper (mg / kg (ppm))	0610	26.50	1.000	25.97	1.803	0.7886	75	0.29	022.43	0
022	Copper (mg / kg (ppm))	0278	26.60	0.8000	25.97	1.803	0.7886	75	0.35	022.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
022	Copper (mg / kg (ppm))	0294	26.60	0.2000	25.97	1.803	0.7886	75	0.35	022.42	0
022	Copper (mg / kg (ppm))	0202	26.81	0.2580	25.97	1.803	0.7886	75	0.46	022.43	0
022	Copper (mg / kg (ppm))	0693	26.88	2.987	25.97	1.803	0.7886	75	0.50	022.42	0
022	Copper (mg / kg (ppm))	0074	27.00	2.000	25.97	1.803	0.7886	75	0.57	022.41	0
022	Copper (mg / kg (ppm))	0164	27.00	0.0000	25.97	1.803	0.7886	75	0.57	022.41	0
022	Copper (mg / kg (ppm))	0035	27.00	0.0000	25.97	1.803	0.7886	75	0.57	022.42	0
022	Copper (mg / kg (ppm))	0035	27.00	0.0000	25.97	1.803	0.7886	75	0.57	022.43	0
022	Copper (mg / kg (ppm))	0083	27.00	0.0000	25.97	1.803	0.7886	75	0.57	022.43	0
022	Copper (mg / kg (ppm))	0049	27.04	0.3900	25.97	1.803	0.7886	75	0.59	022.41	0
022	Copper (mg / kg (ppm))	0357	27.36	0.1220	25.97	1.803	0.7886	75	0.77	022.42	0
022	Copper (mg / kg (ppm))	0148	27.39	0.2200	25.97	1.803	0.7886	75	0.79	022.41	0
022	Copper (mg / kg (ppm))	2248	27.42	1.440	25.97	1.803	0.7886	75	0.80	022.31	0
022	Copper (mg / kg (ppm))	0560	27.50	0.6000	25.97	1.803	0.7886	75	0.85	022.42	0
022	Copper (mg / kg (ppm))	0948	27.53	0.2230	25.97	1.803	0.7886	75	0.86	022.33	0
022	Copper (mg / kg (ppm))	0425	27.60	0.2000	25.97	1.803	0.7886	75	0.90	022.43	0
022	Copper (mg / kg (ppm))	0692	27.62	3.330	25.97	1.803	0.7886	75	0.91	022.42	0
022	Copper (mg / kg (ppm))	0870	27.69	1.690	25.97	1.803	0.7886	75	0.95	022.42	0
022	Copper (mg / kg (ppm))	2196	28.00	0.0000	25.97	1.803	0.7886	75	1.12	022.31	0
022	Copper (mg / kg (ppm))	0511	28.00	0.0000	25.97	1.803	0.7886	75	1.12	022.41	0
022	Copper (mg / kg (ppm))	0297	28.00	0.0000	25.97	1.803	0.7886	75	1.12	022.43	0
022	Copper (mg / kg (ppm))	0590	28.25	1.100	25.97	1.803	0.7886	75	1.26	022.99	0
022	Copper (mg / kg (ppm))	0868	28.50	1.000	25.97	1.803	0.7886	75	1.40	022.31	0
022	Copper (mg / kg (ppm))	0226	28.50	1.000	25.97	1.803	0.7886	75	1.40	022.41	0
022	Copper (mg / kg (ppm))	0505	28.98	0.9200	25.97	1.803	0.7886	75	1.67	022.43	0
022	Copper (mg / kg (ppm))	2188	29.65	2.100	25.97	1.803	0.7886	75	2.04	022.31	0
022	Copper (mg / kg (ppm))	0358	30.53	3.160	25.97	1.803	0.7886	75	2.53	022.41	0
022	Copper (mg / kg (ppm))	0187	35.22	0.1700	25.97	1.803	0.7886	75	5.13	022.42	0
022	Copper (mg / kg (ppm))	0689	45.65	0.5000	25.97	1.803	0.7886	75	10.92	022.31	0
022	Copper (mg / kg (ppm))	0038	54.61	1.340	25.97	1.803	0.7886	75	15.89	022.43	2
025	Iron (mg / kg (ppm))	0536	0.1800	0.0000	325.3	27.35	7.037	76	-11.89	025.31	0
025	Iron (mg / kg (ppm))	0689	75.80	5.200	325.3	27.35	7.037	76	-9.12	025.31	0
025	Iron (mg / kg (ppm))	0154	97.66	0.1600	325.3	27.35	7.037	76	-8.32	025.52	0
025	Iron (mg / kg (ppm))	2053	167.0	2.000	325.3	27.35	7.037	76	-5.79	025.42	0
025	Iron (mg / kg (ppm))	0278	169.0	36.00	325.3	27.35	7.037	76	-5.71	025.42	0
025	Iron (mg / kg (ppm))	2188	214.5	29.60	325.3	27.35	7.037	76	-4.05	025.31	0
025	Iron (mg / kg (ppm))	0726	261.5	15.10	325.3	27.35	7.037	76	-2.33	025.42	0
025	Iron (mg / kg (ppm))	0123	285.0	2.000	325.3	27.35	7.037	76	-1.47	025.41	0
025	Iron (mg / kg (ppm))	0265	290.0	16.00	325.3	27.35	7.037	76	-1.29	025.42	0
025	Iron (mg / kg (ppm))	0345	290.2	3.800	325.3	27.35	7.037	76	-1.28	025.43	0
025	Iron (mg / kg (ppm))	0685	290.5	9.800	325.3	27.35	7.037	76	-1.27	025.42	0
025	Iron (mg / kg (ppm))	0560	297.5	5.200	325.3	27.35	7.037	76	-1.02	025.42	0
025	Iron (mg / kg (ppm))	0598	301.5	1.300	325.3	27.35	7.037	76	-0.87	025.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
025	Iron (mg / kg (ppm))	0169	301.5	3.000	325.3	27.35	7.037	76	-0.87	025.43	0
025	Iron (mg / kg (ppm))	0878	302.5	5.000	325.3	27.35	7.037	76	-0.83	025.41	0
025	Iron (mg / kg (ppm))	0353	302.5	7.000	325.3	27.35	7.037	76	-0.83	025.43	0
025	Iron (mg / kg (ppm))	0190	305.6	10.98	325.3	27.35	7.037	76	-0.72	025.42	0
025	Iron (mg / kg (ppm))	0407	306.4	6.301	325.3	27.35	7.037	76	-0.69	025.41	0
025	Iron (mg / kg (ppm))	0520	307.5	11.00	325.3	27.35	7.037	76	-0.65	025.41	0
025	Iron (mg / kg (ppm))	0720	307.7	0.1000	325.3	27.35	7.037	76	-0.65	025.41	0
025	Iron (mg / kg (ppm))	2248	308.9	0.5200	325.3	27.35	7.037	76	-0.60	025.31	0
025	Iron (mg / kg (ppm))	0692	309.4	8.610	325.3	27.35	7.037	76	-0.58	025.42	0
025	Iron (mg / kg (ppm))	0874	311.0	10.00	325.3	27.35	7.037	76	-0.52	025.31	0
025	Iron (mg / kg (ppm))	2022	311.0	2.000	325.3	27.35	7.037	76	-0.52	025.31	0
025	Iron (mg / kg (ppm))	0511	312.0	2.000	325.3	27.35	7.037	76	-0.49	025.41	0
025	Iron (mg / kg (ppm))	0425	313.7	0.7500	325.3	27.35	7.037	76	-0.42	025.43	0
025	Iron (mg / kg (ppm))	0038	316.5	5.000	325.3	27.35	7.037	76	-0.32	025.31	0
025	Iron (mg / kg (ppm))	0693	318.3	6.134	325.3	27.35	7.037	76	-0.26	025.42	0
025	Iron (mg / kg (ppm))	0164	319.0	2.000	325.3	27.35	7.037	76	-0.23	025.41	0
025	Iron (mg / kg (ppm))	0045	319.5	1.000	325.3	27.35	7.037	76	-0.21	025.42	0
025	Iron (mg / kg (ppm))	0675	319.5	1.990	325.3	27.35	7.037	76	-0.21	025.43	0
025	Iron (mg / kg (ppm))	0098	321.8	5.500	325.3	27.35	7.037	76	-0.13	025.41	0
025	Iron (mg / kg (ppm))	0017	322.1	14.72	325.3	27.35	7.037	76	-0.12	025.43	0
025	Iron (mg / kg (ppm))	0294	325.0	5.120	325.3	27.35	7.037	76	-0.01	025.42	0
025	Iron (mg / kg (ppm))	0619	325.0	10.00	325.3	27.35	7.037	76	-0.01	025.41	0
025	Iron (mg / kg (ppm))	0354	325.3	3.120	325.3	27.35	7.037	76	0.00	025.41	0
025	Iron (mg / kg (ppm))	0100	327.0	12.00	325.3	27.35	7.037	76	0.06	025.99	0
025	Iron (mg / kg (ppm))	0870	327.3	6.300	325.3	27.35	7.037	76	0.07	025.42	0
025	Iron (mg / kg (ppm))	0366	329.5	7.000	325.3	27.35	7.037	76	0.15	025.42	0
025	Iron (mg / kg (ppm))	0910	330.0	12.00	325.3	27.35	7.037	76	0.17	025.41	0
025	Iron (mg / kg (ppm))	0964	330.4	10.01	325.3	27.35	7.037	76	0.18	025.43	0
025	Iron (mg / kg (ppm))	0939	330.9	11.61	325.3	27.35	7.037	76	0.20	025.31	0
025	Iron (mg / kg (ppm))	0505	331.5	10.90	325.3	27.35	7.037	76	0.22	025.31	0
025	Iron (mg / kg (ppm))	0629	332.0	10.00	325.3	27.35	7.037	76	0.24	025.43	0
025	Iron (mg / kg (ppm))	0098	332.4	12.30	325.3	27.35	7.037	76	0.26	025.43	0
025	Iron (mg / kg (ppm))	0242	332.5	3.000	325.3	27.35	7.037	76	0.26	025.99	0
025	Iron (mg / kg (ppm))	0297	333.5	3.000	325.3	27.35	7.037	76	0.30	025.43	0
025	Iron (mg / kg (ppm))	0868	334.0	4.000	325.3	27.35	7.037	76	0.32	025.31	0
025	Iron (mg / kg (ppm))	0407	334.4	10.38	325.3	27.35	7.037	76	0.33	025.43	0
025	Iron (mg / kg (ppm))	0019	335.5	0.5000	325.3	27.35	7.037	76	0.37	025.41	0
025	Iron (mg / kg (ppm))	0510	335.5	1.000	325.3	27.35	7.037	76	0.37	025.43	0
025	Iron (mg / kg (ppm))	0610	336.5	25.00	325.3	27.35	7.037	76	0.41	025.43	0
025	Iron (mg / kg (ppm))	0918	338.1	23.38	325.3	27.35	7.037	76	0.47	025.43	0
025	Iron (mg / kg (ppm))	0171	338.6	0.8000	325.3	27.35	7.037	76	0.49	025.41	0
025	Iron (mg / kg (ppm))	0629	339.0	8.000	325.3	27.35	7.037	76	0.50	025.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
025	Iron (mg / kg (ppm))	0553	339.0	4.000	325.3	27.35	7.037	76	0.50	025.53	0
025	Iron (mg / kg (ppm))	0083	341.5	3.000	325.3	27.35	7.037	76	0.59	025.43	0
025	Iron (mg / kg (ppm))	0939	342.6	3.230	325.3	27.35	7.037	76	0.63	025.53	0
025	Iron (mg / kg (ppm))	0066	343.6	3.000	325.3	27.35	7.037	76	0.67	025.41	0
025	Iron (mg / kg (ppm))	0148	344.2	0.7000	325.3	27.35	7.037	76	0.69	025.41	0
025	Iron (mg / kg (ppm))	2089	345.1	12.31	325.3	27.35	7.037	76	0.73	025.43	0
025	Iron (mg / kg (ppm))	2196	349.7	0.0000	325.3	27.35	7.037	76	0.89	025.31	0
025	Iron (mg / kg (ppm))	0035	350.0	4.000	325.3	27.35	7.037	76	0.90	025.42	0
025	Iron (mg / kg (ppm))	0074	350.5	5.000	325.3	27.35	7.037	76	0.92	025.41	0
025	Iron (mg / kg (ppm))	0049	352.7	6.940	325.3	27.35	7.037	76	1.00	025.41	0
025	Iron (mg / kg (ppm))	0358	353.7	4.210	325.3	27.35	7.037	76	1.04	025.41	0
025	Iron (mg / kg (ppm))	0226	354.0	24.00	325.3	27.35	7.037	76	1.05	025.41	0
025	Iron (mg / kg (ppm))	0968	358.0	20.00	325.3	27.35	7.037	76	1.20	025.43	0
025	Iron (mg / kg (ppm))	0229	358.4	2.800	325.3	27.35	7.037	76	1.21	025.41	0
025	Iron (mg / kg (ppm))	0682	362.4	0.0000	325.3	27.35	7.037	76	1.36	025.43	0
025	Iron (mg / kg (ppm))	0904	362.5	2.990	325.3	27.35	7.037	76	1.36	025.31	0
025	Iron (mg / kg (ppm))	0187	365.7	0.4500	325.3	27.35	7.037	76	1.48	025.42	0
025	Iron (mg / kg (ppm))	0037	372.9	10.00	325.3	27.35	7.037	76	1.74	025.42	0
025	Iron (mg / kg (ppm))	0035	387.5	1.000	325.3	27.35	7.037	76	2.27	025.43	0
025	Iron (mg / kg (ppm))	2246	400.0	0.0000	325.3	27.35	7.037	76	2.73	025.31	0
025	Iron (mg / kg (ppm))	0175	436.0	8.000	325.3	27.35	7.037	76	4.05	025.31	0
025	Iron (mg / kg (ppm))	0032	337.0	100.0	325.3	27.35	7.037	76	0.43	025.43	1
025	Iron (mg / kg (ppm))	0042	340.0	78.00	325.3	27.35	7.037	76	0.54	025.42	1
025	Iron (mg / kg (ppm))	2129	348.5	99.00	325.3	27.35	7.037	76	0.85	025.99	1
027	Magnesium (%)	0536	0.0740	0.0020	0.2689	0.0149	0.0057	94	-13.06	027.31	0
027	Magnesium (%)	0889	0.2083	0.0000	0.2689	0.0149	0.0057	94	-4.06	027.99	0
027	Magnesium (%)	2129	0.2209	0.0042	0.2689	0.0149	0.0057	94	-3.22	027.42	0
027	Magnesium (%)	0047	0.2210	0.0080	0.2689	0.0149	0.0057	94	-3.21	027.52	0
027	Magnesium (%)	0510	0.2350	0.0100	0.2689	0.0149	0.0057	94	-2.27	027.43	0
027	Magnesium (%)	0026	0.2360	0.0080	0.2689	0.0149	0.0057	94	-2.20	027.42	0
027	Magnesium (%)	0154	0.2415	0.0176	0.2689	0.0149	0.0057	94	-1.84	027.52	0
027	Magnesium (%)	0407	0.2444	0.0031	0.2689	0.0149	0.0057	94	-1.65	027.41	0
027	Magnesium (%)	0407	0.2445	0.0053	0.2689	0.0149	0.0057	94	-1.64	027.43	0
027	Magnesium (%)	0202	0.2485	0.0010	0.2689	0.0149	0.0057	94	-1.37	027.43	0
027	Magnesium (%)	0520	0.2500	0.0000	0.2689	0.0149	0.0057	94	-1.27	027.41	0
027	Magnesium (%)	0265	0.2500	0.0000	0.2689	0.0149	0.0057	94	-1.27	027.42	0
027	Magnesium (%)	0045	0.2520	0.0000	0.2689	0.0149	0.0057	94	-1.13	027.42	0
027	Magnesium (%)	0003	0.2550	0.0100	0.2689	0.0149	0.0057	94	-0.93	027.41	0
027	Magnesium (%)	0098	0.2550	0.0100	0.2689	0.0149	0.0057	94	-0.93	027.41	0
027	Magnesium (%)	2192	0.2550	0.0100	0.2689	0.0149	0.0057	94	-0.93	027.43	0
027	Magnesium (%)	0229	0.2585	0.0010	0.2689	0.0149	0.0057	94	-0.70	027.41	0
027	Magnesium (%)	0175	0.2600	0.0000	0.2689	0.0149	0.0057	94	-0.60	027.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
027	Magnesium (%)	0629	0.2600	0.0000	0.2689	0.0149	0.0057	94	-0.60	027.31	0
027	Magnesium (%)	0689	0.2600	0.0000	0.2689	0.0149	0.0057	94	-0.60	027.31	0
027	Magnesium (%)	0004	0.2600	0.0000	0.2689	0.0149	0.0057	94	-0.60	027.41	0
027	Magnesium (%)	0019	0.2600	0.0000	0.2689	0.0149	0.0057	94	-0.60	027.41	0
027	Magnesium (%)	0148	0.2600	0.0000	0.2689	0.0149	0.0057	94	-0.60	027.41	0
027	Magnesium (%)	0878	0.2600	0.0000	0.2689	0.0149	0.0057	94	-0.60	027.41	0
027	Magnesium (%)	0035	0.2600	0.0000	0.2689	0.0149	0.0057	94	-0.60	027.42	0
027	Magnesium (%)	0366	0.2600	0.0200	0.2689	0.0149	0.0057	94	-0.60	027.42	0
027	Magnesium (%)	0685	0.2600	0.0000	0.2689	0.0149	0.0057	94	-0.60	027.42	0
027	Magnesium (%)	0035	0.2600	0.0000	0.2689	0.0149	0.0057	94	-0.60	027.43	0
027	Magnesium (%)	0083	0.2600	0.0000	0.2689	0.0149	0.0057	94	-0.60	027.43	0
027	Magnesium (%)	0037	0.2635	0.0010	0.2689	0.0149	0.0057	94	-0.36	027.42	0
027	Magnesium (%)	0695	0.2640	0.0040	0.2689	0.0149	0.0057	94	-0.33	027.41	0
027	Magnesium (%)	0345	0.2640	0.0020	0.2689	0.0149	0.0057	94	-0.33	027.43	0
027	Magnesium (%)	0142	0.2645	0.0022	0.2689	0.0149	0.0057	94	-0.30	027.31	0
027	Magnesium (%)	0038	0.2649	0.0109	0.2689	0.0149	0.0057	94	-0.27	027.31	0
027	Magnesium (%)	0169	0.2650	0.0100	0.2689	0.0149	0.0057	94	-0.26	027.32	0
027	Magnesium (%)	0049	0.2650	0.0100	0.2689	0.0149	0.0057	94	-0.26	027.41	0
027	Magnesium (%)	0354	0.2650	0.0100	0.2689	0.0149	0.0057	94	-0.26	027.41	0
027	Magnesium (%)	0511	0.2650	0.0100	0.2689	0.0149	0.0057	94	-0.26	027.41	0
027	Magnesium (%)	0692	0.2650	0.0100	0.2689	0.0149	0.0057	94	-0.26	027.42	0
027	Magnesium (%)	0353	0.2650	0.0100	0.2689	0.0149	0.0057	94	-0.26	027.43	0
027	Magnesium (%)	0505	0.2650	0.0100	0.2689	0.0149	0.0057	94	-0.26	027.43	0
027	Magnesium (%)	0590	0.2650	0.0100	0.2689	0.0149	0.0057	94	-0.26	027.99	0
027	Magnesium (%)	0870	0.2663	0.0077	0.2689	0.0149	0.0057	94	-0.18	027.42	0
027	Magnesium (%)	0017	0.2665	0.0050	0.2689	0.0149	0.0057	94	-0.16	027.43	0
027	Magnesium (%)	0656	0.2700	0.0400	0.2689	0.0149	0.0057	94	0.07	027.31	0
027	Magnesium (%)	0868	0.2700	0.0000	0.2689	0.0149	0.0057	94	0.07	027.31	0
027	Magnesium (%)	0720	0.2700	0.0000	0.2689	0.0149	0.0057	94	0.07	027.41	0
027	Magnesium (%)	0910	0.2700	0.0000	0.2689	0.0149	0.0057	94	0.07	027.41	0
027	Magnesium (%)	0294	0.2700	0.0000	0.2689	0.0149	0.0057	94	0.07	027.42	0
027	Magnesium (%)	2053	0.2700	0.0200	0.2689	0.0149	0.0057	94	0.07	027.42	0
027	Magnesium (%)	0032	0.2700	0.0200	0.2689	0.0149	0.0057	94	0.07	027.43	0
027	Magnesium (%)	0089	0.2700	0.0000	0.2689	0.0149	0.0057	94	0.07	027.43	0
027	Magnesium (%)	0297	0.2700	0.0000	0.2689	0.0149	0.0057	94	0.07	027.43	0
027	Magnesium (%)	0425	0.2700	0.0000	0.2689	0.0149	0.0057	94	0.07	027.43	0
027	Magnesium (%)	0610	0.2700	0.0000	0.2689	0.0149	0.0057	94	0.07	027.43	0
027	Magnesium (%)	0675	0.2700	0.0000	0.2689	0.0149	0.0057	94	0.07	027.43	0
027	Magnesium (%)	0242	0.2700	0.0000	0.2689	0.0149	0.0057	94	0.07	027.99	0
027	Magnesium (%)	0357	0.2704	0.0011	0.2689	0.0149	0.0057	94	0.10	027.42	0
027	Magnesium (%)	0529	0.2705	0.0090	0.2689	0.0149	0.0057	94	0.11	027.31	0
027	Magnesium (%)	0598	0.2714	0.0032	0.2689	0.0149	0.0057	94	0.17	027.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
027	Magnesium (%)	0964	0.2715	0.0098	0.2689	0.0149	0.0057	94	0.17	027.43	0
027	Magnesium (%)	0968	0.2720	0.0000	0.2689	0.0149	0.0057	94	0.21	027.43	0
027	Magnesium (%)	0560	0.2722	0.0080	0.2689	0.0149	0.0057	94	0.22	027.42	0
027	Magnesium (%)	0874	0.2750	0.0040	0.2689	0.0149	0.0057	94	0.41	027.31	0
027	Magnesium (%)	0164	0.2750	0.0140	0.2689	0.0149	0.0057	94	0.41	027.41	0
027	Magnesium (%)	0190	0.2750	0.0100	0.2689	0.0149	0.0057	94	0.41	027.42	0
027	Magnesium (%)	0098	0.2750	0.0100	0.2689	0.0149	0.0057	94	0.41	027.43	0
027	Magnesium (%)	2248	0.2765	0.0150	0.2689	0.0149	0.0057	94	0.51	027.31	0
027	Magnesium (%)	0074	0.2765	0.0010	0.2689	0.0149	0.0057	94	0.51	027.41	0
027	Magnesium (%)	0726	0.2774	0.0004	0.2689	0.0149	0.0057	94	0.57	027.42	0
027	Magnesium (%)	0171	0.2780	0.0000	0.2689	0.0149	0.0057	94	0.61	027.41	0
027	Magnesium (%)	2196	0.2800	0.0000	0.2689	0.0149	0.0057	94	0.74	027.31	0
027	Magnesium (%)	2246	0.2800	0.0000	0.2689	0.0149	0.0057	94	0.74	027.31	0
027	Magnesium (%)	0042	0.2800	0.0160	0.2689	0.0149	0.0057	94	0.74	027.42	0
027	Magnesium (%)	0278	0.2800	0.0000	0.2689	0.0149	0.0057	94	0.74	027.42	0
027	Magnesium (%)	0629	0.2800	0.0200	0.2689	0.0149	0.0057	94	0.74	027.43	0
027	Magnesium (%)	2022	0.2810	0.0040	0.2689	0.0149	0.0057	94	0.81	027.31	0
027	Magnesium (%)	0027	0.2810	0.0140	0.2689	0.0149	0.0057	94	0.81	027.43	0
027	Magnesium (%)	0918	0.2845	0.0030	0.2689	0.0149	0.0057	94	1.04	027.43	0
027	Magnesium (%)	0553	0.2845	0.0070	0.2689	0.0149	0.0057	94	1.04	027.53	0
027	Magnesium (%)	0619	0.2860	0.0040	0.2689	0.0149	0.0057	94	1.14	027.41	0
027	Magnesium (%)	0001	0.2865	0.0041	0.2689	0.0149	0.0057	94	1.17	027.31	0
027	Magnesium (%)	0939	0.2900	0.0000	0.2689	0.0149	0.0057	94	1.41	027.31	0
027	Magnesium (%)	0939	0.2900	0.0000	0.2689	0.0149	0.0057	94	1.41	027.53	0
027	Magnesium (%)	0100	0.2900	0.0000	0.2689	0.0149	0.0057	94	1.41	027.99	0
027	Magnesium (%)	0066	0.2910	0.0040	0.2689	0.0149	0.0057	94	1.48	027.41	0
027	Magnesium (%)	0693	0.2915	0.0050	0.2689	0.0149	0.0057	94	1.51	027.42	0
027	Magnesium (%)	0226	0.2950	0.0100	0.2689	0.0149	0.0057	94	1.75	027.41	0
027	Magnesium (%)	2089	0.2950	0.0300	0.2689	0.0149	0.0057	94	1.75	027.43	0
027	Magnesium (%)	0948	0.2960	0.0020	0.2689	0.0149	0.0057	94	1.81	027.33	0
027	Magnesium (%)	0187	0.2986	0.0027	0.2689	0.0149	0.0057	94	1.99	027.42	0
027	Magnesium (%)	0358	0.3000	0.0200	0.2689	0.0149	0.0057	94	2.08	027.41	0
027	Magnesium (%)	0682	0.3000	0.0000	0.2689	0.0149	0.0057	94	2.08	027.43	0
027	Magnesium (%)	2234	0.3200	0.0000	0.2689	0.0149	0.0057	94	3.42	027.43	0
027	Magnesium (%)	0596	0.3350	0.0900	0.2689	0.0149	0.0057	94	4.43	027.31	1
027	Magnesium (%)	0011	0.7921	0.0000	0.2689	0.0149	0.0057	94	35.04	027.41	2
028	Manganese (mg / kg (ppm))	2188	75.10	21.40	142.0	14.03	4.882	84	-4.77	028.31	0
028	Manganese (mg / kg (ppm))	0619	94.40	6.400	142.0	14.03	4.882	84	-3.39	028.41	0
028	Manganese (mg / kg (ppm))	2246	100.0	0.0000	142.0	14.03	4.882	84	-2.99	028.34	0
028	Manganese (mg / kg (ppm))	0689	104.0	0.1000	142.0	14.03	4.882	84	-2.71	028.31	0
028	Manganese (mg / kg (ppm))	0354	107.9	1.050	142.0	14.03	4.882	84	-2.43	028.41	0
028	Manganese (mg / kg (ppm))	0520	112.0	0.0000	142.0	14.03	4.882	84	-2.14	028.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
028	Manganese (mg / kg (ppm))	0685	121.0	1.490	142.0	14.03	4.882	84	-1.50	028.42	0
028	Manganese (mg / kg (ppm))	0026	121.6	3.309	142.0	14.03	4.882	84	-1.46	028.42	0
028	Manganese (mg / kg (ppm))	0948	122.3	5.132	142.0	14.03	4.882	84	-1.40	028.33	0
028	Manganese (mg / kg (ppm))	0598	123.1	0.9000	142.0	14.03	4.882	84	-1.35	028.41	0
028	Manganese (mg / kg (ppm))	0510	124.0	0.0000	142.0	14.03	4.882	84	-1.28	028.43	0
028	Manganese (mg / kg (ppm))	0003	124.5	3.000	142.0	14.03	4.882	84	-1.25	028.41	0
028	Manganese (mg / kg (ppm))	0720	125.4	0.9600	142.0	14.03	4.882	84	-1.18	028.41	0
028	Manganese (mg / kg (ppm))	2129	126.7	4.500	142.0	14.03	4.882	84	-1.10	028.99	0
028	Manganese (mg / kg (ppm))	0878	127.0	12.00	142.0	14.03	4.882	84	-1.07	028.41	0
028	Manganese (mg / kg (ppm))	0171	129.5	1.000	142.0	14.03	4.882	84	-0.89	028.41	0
028	Manganese (mg / kg (ppm))	0407	129.6	2.255	142.0	14.03	4.882	84	-0.89	028.41	0
028	Manganese (mg / kg (ppm))	2022	132.0	0.0000	142.0	14.03	4.882	84	-0.71	028.31	0
028	Manganese (mg / kg (ppm))	0675	132.3	5.400	142.0	14.03	4.882	84	-0.69	028.43	0
028	Manganese (mg / kg (ppm))	0939	133.5	0.7800	142.0	14.03	4.882	84	-0.61	028.31	0
028	Manganese (mg / kg (ppm))	0353	133.5	3.000	142.0	14.03	4.882	84	-0.61	028.43	0
028	Manganese (mg / kg (ppm))	0629	134.5	5.000	142.0	14.03	4.882	84	-0.54	028.43	0
028	Manganese (mg / kg (ppm))	0265	136.0	6.000	142.0	14.03	4.882	84	-0.43	028.42	0
028	Manganese (mg / kg (ppm))	0693	136.8	6.850	142.0	14.03	4.882	84	-0.38	028.42	0
028	Manganese (mg / kg (ppm))	0045	137.0	2.000	142.0	14.03	4.882	84	-0.36	028.42	0
028	Manganese (mg / kg (ppm))	0345	137.0	4.000	142.0	14.03	4.882	84	-0.36	028.43	0
028	Manganese (mg / kg (ppm))	0964	137.4	5.627	142.0	14.03	4.882	84	-0.33	028.43	0
028	Manganese (mg / kg (ppm))	0529	137.5	2.800	142.0	14.03	4.882	84	-0.32	028.31	0
028	Manganese (mg / kg (ppm))	0066	137.9	3.630	142.0	14.03	4.882	84	-0.30	028.44	0
028	Manganese (mg / kg (ppm))	0870	138.2	6.700	142.0	14.03	4.882	84	-0.28	028.42	0
028	Manganese (mg / kg (ppm))	0226	138.5	15.00	142.0	14.03	4.882	84	-0.25	028.41	0
028	Manganese (mg / kg (ppm))	0910	138.5	7.000	142.0	14.03	4.882	84	-0.25	028.41	0
028	Manganese (mg / kg (ppm))	0407	138.7	7.598	142.0	14.03	4.882	84	-0.24	028.43	0
028	Manganese (mg / kg (ppm))	0629	139.0	4.000	142.0	14.03	4.882	84	-0.22	028.31	0
028	Manganese (mg / kg (ppm))	2234	139.4	10.76	142.0	14.03	4.882	84	-0.18	028.43	0
028	Manganese (mg / kg (ppm))	0098	140.4	2.200	142.0	14.03	4.882	84	-0.12	028.41	0
028	Manganese (mg / kg (ppm))	0357	140.8	3.393	142.0	14.03	4.882	84	-0.08	028.42	0
028	Manganese (mg / kg (ppm))	0074	141.5	7.000	142.0	14.03	4.882	84	-0.04	028.41	0
028	Manganese (mg / kg (ppm))	0164	141.5	1.000	142.0	14.03	4.882	84	-0.04	028.41	0
028	Manganese (mg / kg (ppm))	0242	141.5	1.000	142.0	14.03	4.882	84	-0.04	028.99	0
028	Manganese (mg / kg (ppm))	2248	141.7	0.2400	142.0	14.03	4.882	84	-0.02	028.31	0
028	Manganese (mg / kg (ppm))	0610	142.0	16.00	142.0	14.03	4.882	84	0.00	028.43	0
028	Manganese (mg / kg (ppm))	0682	142.2	0.0000	142.0	14.03	4.882	84	0.01	028.43	0
028	Manganese (mg / kg (ppm))	0425	142.4	2.300	142.0	14.03	4.882	84	0.02	028.43	0
028	Manganese (mg / kg (ppm))	0278	142.5	1.000	142.0	14.03	4.882	84	0.03	028.42	0
028	Manganese (mg / kg (ppm))	0968	143.5	1.000	142.0	14.03	4.882	84	0.11	028.43	0
028	Manganese (mg / kg (ppm))	0505	143.7	5.000	142.0	14.03	4.882	84	0.12	028.43	0
028	Manganese (mg / kg (ppm))	2053	144.0	4.000	142.0	14.03	4.882	84	0.14	028.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
028	Manganese (mg / kg (ppm))	0098	144.2	5.800	142.0	14.03	4.882	84	0.16	028.43	0
028	Manganese (mg / kg (ppm))	0083	145.0	2.000	142.0	14.03	4.882	84	0.21	028.43	0
028	Manganese (mg / kg (ppm))	0868	145.5	7.000	142.0	14.03	4.882	84	0.25	028.31	0
028	Manganese (mg / kg (ppm))	0511	146.0	10.00	142.0	14.03	4.882	84	0.28	028.41	0
028	Manganese (mg / kg (ppm))	0190	146.3	6.380	142.0	14.03	4.882	84	0.31	028.42	0
028	Manganese (mg / kg (ppm))	0939	146.3	0.4200	142.0	14.03	4.882	84	0.31	028.53	0
028	Manganese (mg / kg (ppm))	2196	147.8	0.0000	142.0	14.03	4.882	84	0.41	028.31	0
028	Manganese (mg / kg (ppm))	0035	148.0	4.000	142.0	14.03	4.882	84	0.43	028.42	0
028	Manganese (mg / kg (ppm))	0035	148.5	3.000	142.0	14.03	4.882	84	0.46	028.43	0
028	Manganese (mg / kg (ppm))	0229	148.7	0.5000	142.0	14.03	4.882	84	0.47	028.41	0
028	Manganese (mg / kg (ppm))	0726	149.1	0.6000	142.0	14.03	4.882	84	0.50	028.42	0
028	Manganese (mg / kg (ppm))	0560	149.2	11.10	142.0	14.03	4.882	84	0.51	028.42	0
028	Manganese (mg / kg (ppm))	0656	149.6	8.210	142.0	14.03	4.882	84	0.54	028.31	0
028	Manganese (mg / kg (ppm))	0904	150.0	0.4100	142.0	14.03	4.882	84	0.57	028.31	0
028	Manganese (mg / kg (ppm))	0169	150.0	14.00	142.0	14.03	4.882	84	0.57	028.43	0
028	Manganese (mg / kg (ppm))	0553	150.5	15.00	142.0	14.03	4.882	84	0.60	028.53	0
028	Manganese (mg / kg (ppm))	0297	151.0	2.000	142.0	14.03	4.882	84	0.64	028.43	0
028	Manganese (mg / kg (ppm))	0187	151.1	1.470	142.0	14.03	4.882	84	0.64	028.42	0
028	Manganese (mg / kg (ppm))	0037	151.5	0.5000	142.0	14.03	4.882	84	0.67	028.42	0
028	Manganese (mg / kg (ppm))	0017	152.0	16.36	142.0	14.03	4.882	84	0.71	028.43	0
028	Manganese (mg / kg (ppm))	0100	153.0	0.0000	142.0	14.03	4.882	84	0.78	028.99	0
028	Manganese (mg / kg (ppm))	0019	153.1	2.100	142.0	14.03	4.882	84	0.79	028.41	0
028	Manganese (mg / kg (ppm))	0202	153.3	2.736	142.0	14.03	4.882	84	0.80	028.43	0
028	Manganese (mg / kg (ppm))	0692	156.1	3.200	142.0	14.03	4.882	84	1.00	028.42	0
028	Manganese (mg / kg (ppm))	0049	157.8	0.3200	142.0	14.03	4.882	84	1.12	028.41	0
028	Manganese (mg / kg (ppm))	0038	160.0	20.00	142.0	14.03	4.882	84	1.28	028.32	0
028	Manganese (mg / kg (ppm))	0032	160.0	4.000	142.0	14.03	4.882	84	1.28	028.43	0
028	Manganese (mg / kg (ppm))	0590	160.0	16.00	142.0	14.03	4.882	84	1.28	028.99	0
028	Manganese (mg / kg (ppm))	0358	160.2	1.510	142.0	14.03	4.882	84	1.29	028.41	0
028	Manganese (mg / kg (ppm))	0366	162.5	21.00	142.0	14.03	4.882	84	1.46	028.42	0
028	Manganese (mg / kg (ppm))	0148	162.5	8.880	142.0	14.03	4.882	84	1.46	028.41	0
028	Manganese (mg / kg (ppm))	0918	164.4	10.57	142.0	14.03	4.882	84	1.59	028.43	0
028	Manganese (mg / kg (ppm))	0294	165.8	2.820	142.0	14.03	4.882	84	1.69	028.42	0
028	Manganese (mg / kg (ppm))	0941	168.8	0.0000	142.0	14.03	4.882	84	1.91	028.43	0
028	Manganese (mg / kg (ppm))	2089	168.9	4.400	142.0	14.03	4.882	84	1.91	028.43	0
028	Manganese (mg / kg (ppm))	0011	399.8	0.0000	142.0	14.03	4.882	84	18.37	028.41	0
028	Manganese (mg / kg (ppm))	0047	144.1	28.28	142.0	14.03	4.882	84	0.15	028.52	1
028	Manganese (mg / kg (ppm))	0042	159.0	36.00	142.0	14.03	4.882	84	1.21	028.42	1
028	Manganese (mg / kg (ppm))	0536	546.2	2.050	142.0	14.03	4.882	84	28.80	028.31	2
031	Phosphorus (%)	0043	0.3886	0.0000	0.9238	0.0488	0.0227	131	-10.96	031.43	0
031	Phosphorus (%)	0536	0.5750	0.0100	0.9238	0.0488	0.0227	131	-7.14	031.06	0
031	Phosphorus (%)	0889	0.6690	0.0159	0.9238	0.0488	0.0227	131	-5.22	031.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	0123	0.7700	0.0200	0.9238	0.0488	0.0227	131	-3.15	031.41	0
031	Phosphorus (%)	0108	0.7800	0.0600	0.9238	0.0488	0.0227	131	-2.95	031.01	0
031	Phosphorus (%)	0689	0.7950	0.0100	0.9238	0.0488	0.0227	131	-2.64	031.01	0
031	Phosphorus (%)	0693	0.8050	0.0320	0.9238	0.0488	0.0227	131	-2.43	031.42	0
031	Phosphorus (%)	2248	0.8360	0.0200	0.9238	0.0488	0.0227	131	-1.80	031.01	0
031	Phosphorus (%)	0294	0.8450	0.0100	0.9238	0.0488	0.0227	131	-1.61	031.42	0
031	Phosphorus (%)	2181	0.8500	0.0000	0.9238	0.0488	0.0227	131	-1.51	031.01	0
031	Phosphorus (%)	0687	0.8500	0.1000	0.9238	0.0488	0.0227	131	-1.51	031.01	0
031	Phosphorus (%)	2128	0.8515	0.0550	0.9238	0.0488	0.0227	131	-1.48	031.01	0
031	Phosphorus (%)	0019	0.8550	0.0100	0.9238	0.0488	0.0227	131	-1.41	031.41	0
031	Phosphorus (%)	0848	0.8550	0.0100	0.9238	0.0488	0.0227	131	-1.41	031.41	0
031	Phosphorus (%)	0939	0.8550	0.0100	0.9238	0.0488	0.0227	131	-1.41	031.53	0
031	Phosphorus (%)	0154	0.8568	0.0110	0.9238	0.0488	0.0227	131	-1.37	031.52	0
031	Phosphorus (%)	0948	0.8575	0.0310	0.9238	0.0488	0.0227	131	-1.36	031.01	0
031	Phosphorus (%)	0144	0.8650	0.0100	0.9238	0.0488	0.0227	131	-1.20	031.41	0
031	Phosphorus (%)	2018	0.8700	0.0200	0.9238	0.0488	0.0227	131	-1.10	031.01	0
031	Phosphorus (%)	0520	0.8700	0.0200	0.9238	0.0488	0.0227	131	-1.10	031.41	0
031	Phosphorus (%)	0511	0.8750	0.0500	0.9238	0.0488	0.0227	131	-1.00	031.01	0
031	Phosphorus (%)	0035	0.8750	0.0100	0.9238	0.0488	0.0227	131	-1.00	031.42	0
031	Phosphorus (%)	0142	0.8755	0.0005	0.9238	0.0488	0.0227	131	-0.99	031.01	0
031	Phosphorus (%)	0089	0.8800	0.0000	0.9238	0.0488	0.0227	131	-0.90	031.43	0
031	Phosphorus (%)	0175	0.8850	0.0300	0.9238	0.0488	0.0227	131	-0.79	031.01	0
031	Phosphorus (%)	2246	0.8850	0.0100	0.9238	0.0488	0.0227	131	-0.79	031.01	0
031	Phosphorus (%)	0004	0.8850	0.0100	0.9238	0.0488	0.0227	131	-0.79	031.41	0
031	Phosphorus (%)	0353	0.8850	0.0100	0.9238	0.0488	0.0227	131	-0.79	031.43	0
031	Phosphorus (%)	2129	0.8861	0.0719	0.9238	0.0488	0.0227	131	-0.77	031.42	0
031	Phosphorus (%)	2199	0.8900	0.0200	0.9238	0.0488	0.0227	131	-0.69	031.01	0
031	Phosphorus (%)	0298	0.8900	0.0200	0.9238	0.0488	0.0227	131	-0.69	031.41	0
031	Phosphorus (%)	2009	0.8917	0.0005	0.9238	0.0488	0.0227	131	-0.66	031.01	0
031	Phosphorus (%)	0407	0.8944	0.0166	0.9238	0.0488	0.0227	131	-0.60	031.43	0
031	Phosphorus (%)	0529	0.8950	0.0100	0.9238	0.0488	0.0227	131	-0.59	031.01	0
031	Phosphorus (%)	2200	0.8950	0.0300	0.9238	0.0488	0.0227	131	-0.59	031.01	0
031	Phosphorus (%)	0265	0.8950	0.0300	0.9238	0.0488	0.0227	131	-0.59	031.42	0
031	Phosphorus (%)	0874	0.8995	0.0110	0.9238	0.0488	0.0227	131	-0.50	031.01	0
031	Phosphorus (%)	2192	0.9000	0.0000	0.9238	0.0488	0.0227	131	-0.49	031.01	0
031	Phosphorus (%)	0100	0.9000	0.0000	0.9238	0.0488	0.0227	131	-0.49	031.99	0
031	Phosphorus (%)	0164	0.9040	0.0480	0.9238	0.0488	0.0227	131	-0.41	031.41	0
031	Phosphorus (%)	0297	0.9050	0.0300	0.9238	0.0488	0.0227	131	-0.38	031.43	0
031	Phosphorus (%)	0407	0.9059	0.0022	0.9238	0.0488	0.0227	131	-0.37	031.41	0
031	Phosphorus (%)	0560	0.9067	0.0368	0.9238	0.0488	0.0227	131	-0.35	031.42	0
031	Phosphorus (%)	0038	0.9084	0.0450	0.9238	0.0488	0.0227	131	-0.32	031.01	0
031	Phosphorus (%)	0726	0.9099	0.0046	0.9238	0.0488	0.0227	131	-0.28	031.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	0676	0.9100	0.0600	0.9238	0.0488	0.0227	131	-0.28	031.99	0
031	Phosphorus (%)	2196	0.9100	0.0000	0.9238	0.0488	0.0227	131	-0.28	031.01	0
031	Phosphorus (%)	0685	0.9100	0.0000	0.9238	0.0488	0.0227	131	-0.28	031.41	0
031	Phosphorus (%)	0910	0.9100	0.0000	0.9238	0.0488	0.0227	131	-0.28	031.41	0
031	Phosphorus (%)	0884	0.9150	0.0500	0.9238	0.0488	0.0227	131	-0.18	031.01	0
031	Phosphorus (%)	0045	0.9160	0.0060	0.9238	0.0488	0.0227	131	-0.16	031.42	0
031	Phosphorus (%)	0036	0.9179	0.0087	0.9238	0.0488	0.0227	131	-0.12	031.43	0
031	Phosphorus (%)	0590	0.9200	0.0000	0.9238	0.0488	0.0227	131	-0.08	031.01	0
031	Phosphorus (%)	2022	0.9200	0.0000	0.9238	0.0488	0.0227	131	-0.08	031.01	0
031	Phosphorus (%)	0505	0.9200	0.0000	0.9238	0.0488	0.0227	131	-0.08	031.02	0
031	Phosphorus (%)	0169	0.9200	0.0200	0.9238	0.0488	0.0227	131	-0.08	031.03	0
031	Phosphorus (%)	0354	0.9200	0.0200	0.9238	0.0488	0.0227	131	-0.08	031.41	0
031	Phosphorus (%)	0190	0.9200	0.0200	0.9238	0.0488	0.0227	131	-0.08	031.42	0
031	Phosphorus (%)	0010	0.9200	0.0000	0.9238	0.0488	0.0227	131	-0.08	031.43	0
031	Phosphorus (%)	0036	0.9213	0.0014	0.9238	0.0488	0.0227	131	-0.05	031.03	0
031	Phosphorus (%)	0014	0.9245	0.0410	0.9238	0.0488	0.0227	131	0.01	031.42	0
031	Phosphorus (%)	2193	0.9250	0.0100	0.9238	0.0488	0.0227	131	0.03	031.01	0
031	Phosphorus (%)	0171	0.9250	0.0100	0.9238	0.0488	0.0227	131	0.03	031.41	0
031	Phosphorus (%)	0720	0.9250	0.0100	0.9238	0.0488	0.0227	131	0.03	031.41	0
031	Phosphorus (%)	0425	0.9250	0.0100	0.9238	0.0488	0.0227	131	0.03	031.43	0
031	Phosphorus (%)	0918	0.9255	0.0190	0.9238	0.0488	0.0227	131	0.04	031.43	0
031	Phosphorus (%)	0626	0.9265	0.0150	0.9238	0.0488	0.0227	131	0.06	031.01	0
031	Phosphorus (%)	0849	0.9275	0.0050	0.9238	0.0488	0.0227	131	0.08	031.01	0
031	Phosphorus (%)	0897	0.9300	0.0200	0.9238	0.0488	0.0227	131	0.13	031.01	0
031	Phosphorus (%)	0148	0.9300	0.0400	0.9238	0.0488	0.0227	131	0.13	031.41	0
031	Phosphorus (%)	0366	0.9300	0.0600	0.9238	0.0488	0.0227	131	0.13	031.42	0
031	Phosphorus (%)	0345	0.9300	0.0160	0.9238	0.0488	0.0227	131	0.13	031.43	0
031	Phosphorus (%)	0939	0.9300	0.0000	0.9238	0.0488	0.0227	131	0.13	031.01	0
031	Phosphorus (%)	2126	0.9300	0.0000	0.9238	0.0488	0.0227	131	0.13	031.01	0
031	Phosphorus (%)	0049	0.9300	0.0000	0.9238	0.0488	0.0227	131	0.13	031.41	0
031	Phosphorus (%)	0278	0.9300	0.0000	0.9238	0.0488	0.0227	131	0.13	031.42	0
031	Phosphorus (%)	2089	0.9300	0.0000	0.9238	0.0488	0.0227	131	0.13	031.43	0
031	Phosphorus (%)	0695	0.9320	0.0160	0.9238	0.0488	0.0227	131	0.17	031.41	0
031	Phosphorus (%)	0027	0.9325	0.0870	0.9238	0.0488	0.0227	131	0.18	031.43	0
031	Phosphorus (%)	0968	0.9325	0.0010	0.9238	0.0488	0.0227	131	0.18	031.43	0
031	Phosphorus (%)	2190	0.9346	0.0093	0.9238	0.0488	0.0227	131	0.22	031.01	0
031	Phosphorus (%)	0683	0.9350	0.0300	0.9238	0.0488	0.0227	131	0.23	031.01	0
031	Phosphorus (%)	2188	0.9350	0.0100	0.9238	0.0488	0.0227	131	0.23	031.01	0
031	Phosphorus (%)	2053	0.9350	0.0100	0.9238	0.0488	0.0227	131	0.23	031.02	0
031	Phosphorus (%)	0001	0.9350	0.0060	0.9238	0.0488	0.0227	131	0.23	031.03	0
031	Phosphorus (%)	0242	0.9350	0.0100	0.9238	0.0488	0.0227	131	0.23	031.99	0
031	Phosphorus (%)	2062	0.9374	0.0207	0.9238	0.0488	0.0227	131	0.28	031.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	0037	0.9390	0.0160	0.9238	0.0488	0.0227	131	0.31	031.42	0
031	Phosphorus (%)	0964	0.9411	0.0166	0.9238	0.0488	0.0227	131	0.35	031.43	0
031	Phosphorus (%)	0885	0.9414	0.0013	0.9238	0.0488	0.0227	131	0.36	031.01	0
031	Phosphorus (%)	0629	0.9450	0.0100	0.9238	0.0488	0.0227	131	0.43	031.01	0
031	Phosphorus (%)	0229	0.9450	0.0180	0.9238	0.0488	0.0227	131	0.43	031.41	0
031	Phosphorus (%)	0083	0.9450	0.0300	0.9238	0.0488	0.0227	131	0.43	031.43	0
031	Phosphorus (%)	0868	0.9500	0.0200	0.9238	0.0488	0.0227	131	0.54	031.01	0
031	Phosphorus (%)	0878	0.9500	0.0200	0.9238	0.0488	0.0227	131	0.54	031.01	0
031	Phosphorus (%)	2006	0.9510	0.0000	0.9238	0.0488	0.0227	131	0.56	031.01	0
031	Phosphorus (%)	0202	0.9520	0.0220	0.9238	0.0488	0.0227	131	0.58	031.43	0
031	Phosphorus (%)	0553	0.9530	0.0600	0.9238	0.0488	0.0227	131	0.60	031.53	0
031	Phosphorus (%)	0881	0.9535	0.0230	0.9238	0.0488	0.0227	131	0.61	031.01	0
031	Phosphorus (%)	0675	0.9550	0.0100	0.9238	0.0488	0.0227	131	0.64	031.43	0
031	Phosphorus (%)	0066	0.9575	0.0290	0.9238	0.0488	0.0227	131	0.69	031.44	0
031	Phosphorus (%)	0187	0.9580	0.0020	0.9238	0.0488	0.0227	131	0.70	031.42	0
031	Phosphorus (%)	2251	0.9600	0.0400	0.9238	0.0488	0.0227	131	0.74	031.01	0
031	Phosphorus (%)	0619	0.9615	0.0510	0.9238	0.0488	0.0227	131	0.77	031.41	0
031	Phosphorus (%)	0893	0.9650	0.0300	0.9238	0.0488	0.0227	131	0.84	031.00	0
031	Phosphorus (%)	0610	0.9650	0.0100	0.9238	0.0488	0.0227	131	0.84	031.43	0
031	Phosphorus (%)	2109	0.9680	0.0380	0.9238	0.0488	0.0227	131	0.91	031.01	0
031	Phosphorus (%)	0656	0.9700	0.0400	0.9238	0.0488	0.0227	131	0.95	031.01	0
031	Phosphorus (%)	0047	0.9700	0.0800	0.9238	0.0488	0.0227	131	0.95	031.42	0
031	Phosphorus (%)	0692	0.9700	0.1200	0.9238	0.0488	0.0227	131	0.95	031.42	0
031	Phosphorus (%)	0098	0.9714	0.0190	0.9238	0.0488	0.0227	131	0.98	031.41	0
031	Phosphorus (%)	0003	0.9750	0.0100	0.9238	0.0488	0.0227	131	1.05	031.41	0
031	Phosphorus (%)	0629	0.9750	0.0100	0.9238	0.0488	0.0227	131	1.05	031.43	0
031	Phosphorus (%)	0098	0.9783	0.0082	0.9238	0.0488	0.0227	131	1.12	031.43	0
031	Phosphorus (%)	0682	0.9800	0.0000	0.9238	0.0488	0.0227	131	1.15	031.43	0
031	Phosphorus (%)	0042	0.9825	0.0750	0.9238	0.0488	0.0227	131	1.20	031.42	0
031	Phosphorus (%)	2198	0.9850	0.0100	0.9238	0.0488	0.0227	131	1.25	031.01	0
031	Phosphorus (%)	0357	0.9872	0.0019	0.9238	0.0488	0.0227	131	1.30	031.42	0
031	Phosphorus (%)	0226	0.9900	0.0200	0.9238	0.0488	0.0227	131	1.36	031.41	0
031	Phosphorus (%)	0035	0.9900	0.0200	0.9238	0.0488	0.0227	131	1.36	031.43	0
031	Phosphorus (%)	0852	0.9900	0.0200	0.9238	0.0488	0.0227	131	1.36	031.99	0
031	Phosphorus (%)	2259	0.9925	0.0190	0.9238	0.0488	0.0227	131	1.41	031.41	0
031	Phosphorus (%)	2259	0.9950	0.1280	0.9238	0.0488	0.0227	131	1.46	031.42	0
031	Phosphorus (%)	0598	0.9962	0.0197	0.9238	0.0488	0.0227	131	1.48	031.41	0
031	Phosphorus (%)	0870	0.9980	0.0540	0.9238	0.0488	0.0227	131	1.52	031.42	0
031	Phosphorus (%)	0074	1.000	0.0200	0.9238	0.0488	0.0227	131	1.56	031.41	0
031	Phosphorus (%)	0017	1.005	0.0780	0.9238	0.0488	0.0227	131	1.66	031.43	0
031	Phosphorus (%)	0032	1.020	0.0200	0.9238	0.0488	0.0227	131	1.97	031.43	0
031	Phosphorus (%)	0358	1.025	0.1300	0.9238	0.0488	0.0227	131	2.07	031.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	0510	1.050	0.0000	0.9238	0.0488	0.0227	131	2.59	031.43	0
031	Phosphorus (%)	2234	1.275	0.0300	0.9238	0.0488	0.0227	131	7.19	031.43	0
031	Phosphorus (%)	0026	0.7315	0.3410	0.9238	0.0488	0.0227	131	-3.94	031.42	1
031	Phosphorus (%)	0011	3.048	0.0000	0.9238	0.0488	0.0227	131	43.52	031.41	2
032	Potassium (%)	0889	0.0410	0.0158	0.9605	0.0609	0.0228	96	-15.10	032.99	0
032	Potassium (%)	0685	0.5600	0.0400	0.9605	0.0609	0.0228	96	-6.58	032.42	0
032	Potassium (%)	0154	0.7221	0.0542	0.9605	0.0609	0.0228	96	-3.91	032.52	0
032	Potassium (%)	0123	0.7950	0.0300	0.9605	0.0609	0.0228	96	-2.72	032.41	0
032	Potassium (%)	2246	0.8200	0.0000	0.9605	0.0609	0.0228	96	-2.31	032.32	0
032	Potassium (%)	0026	0.8520	0.0100	0.9605	0.0609	0.0228	96	-1.78	032.42	0
032	Potassium (%)	2129	0.8528	0.0334	0.9605	0.0609	0.0228	96	-1.77	032.42	0
032	Potassium (%)	0004	0.8600	0.0600	0.9605	0.0609	0.0228	96	-1.65	032.41	0
032	Potassium (%)	0019	0.8600	0.0000	0.9605	0.0609	0.0228	96	-1.65	032.41	0
032	Potassium (%)	0868	0.8750	0.0300	0.9605	0.0609	0.0228	96	-1.40	032.31	0
032	Potassium (%)	0049	0.8750	0.0500	0.9605	0.0609	0.0228	96	-1.40	032.41	0
032	Potassium (%)	2097	0.8785	0.0330	0.9605	0.0609	0.0228	96	-1.35	032.31	0
032	Potassium (%)	0003	0.8800	0.0200	0.9605	0.0609	0.0228	96	-1.32	032.41	0
032	Potassium (%)	0964	0.8901	0.0147	0.9605	0.0609	0.0228	96	-1.16	032.43	0
032	Potassium (%)	0720	0.8950	0.0100	0.9605	0.0609	0.0228	96	-1.08	032.41	0
032	Potassium (%)	2053	0.9000	0.0000	0.9605	0.0609	0.0228	96	-0.99	032.31	0
032	Potassium (%)	0354	0.9000	0.0200	0.9605	0.0609	0.0228	96	-0.99	032.41	0
032	Potassium (%)	0047	0.9000	0.0600	0.9605	0.0609	0.0228	96	-0.99	032.42	0
032	Potassium (%)	0520	0.9050	0.0100	0.9605	0.0609	0.0228	96	-0.91	032.41	0
032	Potassium (%)	0619	0.9050	0.0200	0.9605	0.0609	0.0228	96	-0.91	032.41	0
032	Potassium (%)	0037	0.9050	0.0320	0.9605	0.0609	0.0228	96	-0.91	032.42	0
032	Potassium (%)	0511	0.9100	0.0400	0.9605	0.0609	0.0228	96	-0.83	032.41	0
032	Potassium (%)	0510	0.9100	0.0000	0.9605	0.0609	0.0228	96	-0.83	032.43	0
032	Potassium (%)	0038	0.9219	0.0177	0.9605	0.0609	0.0228	96	-0.63	032.31	0
032	Potassium (%)	0144	0.9250	0.0100	0.9605	0.0609	0.0228	96	-0.58	032.41	0
032	Potassium (%)	0536	0.9300	0.0200	0.9605	0.0609	0.0228	96	-0.50	032.31	0
032	Potassium (%)	0032	0.9300	0.0600	0.9605	0.0609	0.0228	96	-0.50	032.43	0
032	Potassium (%)	0629	0.9350	0.0100	0.9605	0.0609	0.0228	96	-0.42	032.31	0
032	Potassium (%)	0089	0.9350	0.0100	0.9605	0.0609	0.0228	96	-0.42	032.43	0
032	Potassium (%)	0598	0.9366	0.0129	0.9605	0.0609	0.0228	96	-0.39	032.41	0
032	Potassium (%)	0014	0.9375	0.0090	0.9605	0.0609	0.0228	96	-0.38	032.31	0
032	Potassium (%)	0229	0.9400	0.0080	0.9605	0.0609	0.0228	96	-0.34	032.41	0
032	Potassium (%)	0939	0.9400	0.0000	0.9605	0.0609	0.0228	96	-0.34	032.53	0
032	Potassium (%)	0590	0.9400	0.0800	0.9605	0.0609	0.0228	96	-0.34	032.99	0
032	Potassium (%)	0035	0.9450	0.0100	0.9605	0.0609	0.0228	96	-0.25	032.42	0
032	Potassium (%)	0074	0.9500	0.0200	0.9605	0.0609	0.0228	96	-0.17	032.41	0
032	Potassium (%)	0265	0.9500	0.0200	0.9605	0.0609	0.0228	96	-0.17	032.42	0
032	Potassium (%)	0294	0.9500	0.0000	0.9605	0.0609	0.0228	96	-0.17	032.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
032	Potassium (%)	0353	0.9500	0.0200	0.9605	0.0609	0.0228	96	-0.17	032.43	0
032	Potassium (%)	0164	0.9540	0.0300	0.9605	0.0609	0.0228	96	-0.11	032.41	0
032	Potassium (%)	0529	0.9550	0.0300	0.9605	0.0609	0.0228	96	-0.09	032.31	0
032	Potassium (%)	0098	0.9550	0.0100	0.9605	0.0609	0.0228	96	-0.09	032.43	0
032	Potassium (%)	0278	0.9600	0.0200	0.9605	0.0609	0.0228	96	-0.01	032.42	0
032	Potassium (%)	0948	0.9615	0.0090	0.9605	0.0609	0.0228	96	0.02	032.31	0
032	Potassium (%)	0066	0.9630	0.0080	0.9605	0.0609	0.0228	96	0.04	032.41	0
032	Potassium (%)	0001	0.9630	0.0140	0.9605	0.0609	0.0228	96	0.04	032.99	0
032	Potassium (%)	0695	0.9640	0.0240	0.9605	0.0609	0.0228	96	0.06	032.41	0
032	Potassium (%)	0726	0.9646	0.0004	0.9605	0.0609	0.0228	96	0.07	032.42	0
032	Potassium (%)	0148	0.9650	0.0100	0.9605	0.0609	0.0228	96	0.07	032.41	0
032	Potassium (%)	0010	0.9650	0.0100	0.9605	0.0609	0.0228	96	0.07	032.43	0
032	Potassium (%)	0693	0.9695	0.0250	0.9605	0.0609	0.0228	96	0.15	032.42	0
032	Potassium (%)	0629	0.9700	0.0200	0.9605	0.0609	0.0228	96	0.16	032.43	0
032	Potassium (%)	0100	0.9700	0.0000	0.9605	0.0609	0.0228	96	0.16	032.99	0
032	Potassium (%)	0345	0.9710	0.0100	0.9605	0.0609	0.0228	96	0.17	032.43	0
032	Potassium (%)	0560	0.9734	0.0079	0.9605	0.0609	0.0228	96	0.21	032.42	0
032	Potassium (%)	0171	0.9750	0.0100	0.9605	0.0609	0.0228	96	0.24	032.41	0
032	Potassium (%)	0910	0.9750	0.0100	0.9605	0.0609	0.0228	96	0.24	032.41	0
032	Potassium (%)	0242	0.9750	0.0300	0.9605	0.0609	0.0228	96	0.24	032.99	0
032	Potassium (%)	0202	0.9790	0.0020	0.9605	0.0609	0.0228	96	0.30	032.43	0
032	Potassium (%)	0610	0.9800	0.0000	0.9605	0.0609	0.0228	96	0.32	032.43	0
032	Potassium (%)	0682	0.9800	0.0000	0.9605	0.0609	0.0228	96	0.32	032.43	0
032	Potassium (%)	0036	0.9873	0.0111	0.9605	0.0609	0.0228	96	0.44	032.43	0
032	Potassium (%)	0027	0.9880	0.0760	0.9605	0.0609	0.0228	96	0.45	032.43	0
032	Potassium (%)	0042	0.9895	0.0410	0.9605	0.0609	0.0228	96	0.48	032.42	0
032	Potassium (%)	0297	0.9900	0.0200	0.9605	0.0609	0.0228	96	0.48	032.43	0
032	Potassium (%)	0407	0.9913	0.0083	0.9605	0.0609	0.0228	96	0.51	032.41	0
032	Potassium (%)	0357	0.9913	0.0162	0.9605	0.0609	0.0228	96	0.51	032.42	0
032	Potassium (%)	0553	0.9950	0.0100	0.9605	0.0609	0.0228	96	0.57	032.53	0
032	Potassium (%)	0870	0.9982	0.0377	0.9605	0.0609	0.0228	96	0.62	032.42	0
032	Potassium (%)	0918	0.9995	0.0170	0.9605	0.0609	0.0228	96	0.64	032.43	0
032	Potassium (%)	0190	1.000	0.0200	0.9605	0.0609	0.0228	96	0.65	032.42	0
032	Potassium (%)	0083	1.000	0.0200	0.9605	0.0609	0.0228	96	0.65	032.43	0
032	Potassium (%)	0505	1.000	0.1600	0.9605	0.0609	0.0228	96	0.65	032.43	0
032	Potassium (%)	0675	1.000	0.0000	0.9605	0.0609	0.0228	96	0.65	032.43	0
032	Potassium (%)	0407	1.001	0.0115	0.9605	0.0609	0.0228	96	0.66	032.43	0
032	Potassium (%)	0939	1.005	0.0100	0.9605	0.0609	0.0228	96	0.73	032.31	0
032	Potassium (%)	0098	1.005	0.0300	0.9605	0.0609	0.0228	96	0.73	032.41	0
032	Potassium (%)	0878	1.005	0.0100	0.9605	0.0609	0.0228	96	0.73	032.41	0
032	Potassium (%)	2089	1.005	0.0300	0.9605	0.0609	0.0228	96	0.73	032.43	0
032	Potassium (%)	0656	1.010	0.0400	0.9605	0.0609	0.0228	96	0.81	032.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
032	Potassium (%)	0035	1.010	0.0000	0.9605	0.0609	0.0228	96	0.81	032.43	0
032	Potassium (%)	0425	1.015	0.0100	0.9605	0.0609	0.0228	96	0.90	032.43	0
032	Potassium (%)	0226	1.020	0.0400	0.9605	0.0609	0.0228	96	0.98	032.41	0
032	Potassium (%)	0358	1.020	0.1400	0.9605	0.0609	0.0228	96	0.98	032.41	0
032	Potassium (%)	0187	1.020	0.0000	0.9605	0.0609	0.0228	96	0.98	032.42	0
032	Potassium (%)	0692	1.020	0.0000	0.9605	0.0609	0.0228	96	0.98	032.42	0
032	Potassium (%)	0169	1.030	0.0200	0.9605	0.0609	0.0228	96	1.14	032.32	0
032	Potassium (%)	0142	1.041	0.0197	0.9605	0.0609	0.0228	96	1.33	032.31	0
032	Potassium (%)	0017	1.063	0.0740	0.9605	0.0609	0.0228	96	1.68	032.43	0
032	Potassium (%)	0683	1.070	0.0400	0.9605	0.0609	0.0228	96	1.80	032.31	0
032	Potassium (%)	0045	1.115	0.0300	0.9605	0.0609	0.0228	96	2.54	032.42	0
032	Potassium (%)	0366	1.135	0.0700	0.9605	0.0609	0.0228	96	2.87	032.42	0
032	Potassium (%)	0968	1.137	0.0030	0.9605	0.0609	0.0228	96	2.89	032.43	0
032	Potassium (%)	0874	1.150	0.0000	0.9605	0.0609	0.0228	96	3.11	032.31	0
032	Potassium (%)	2234	1.230	0.0000	0.9605	0.0609	0.0228	96	4.42	032.43	0
032	Potassium (%)	2248	2.580	0.0000	0.9605	0.0609	0.0228	96	26.59	032.02	0
032	Potassium (%)	0011	14,023	28,041	0.9605	0.0609	0.0228	96	230218.75	032.41	2
033	Salt as chloride (%)	2129	0.3075	0.0050	0.4914	0.0463	0.0180	72	-3.98	033.05	0
033	Salt as chloride (%)	2073	0.3300	0.0000	0.4914	0.0463	0.0180	72	-3.49	033.99	0
033	Salt as chloride (%)	0675	0.3500	0.0600	0.4914	0.0463	0.0180	72	-3.06	033.00	0
033	Salt as chloride (%)	0171	0.3600	0.0200	0.4914	0.0463	0.0180	72	-2.84	033.00	0
033	Salt as chloride (%)	0298	0.3800	0.0200	0.4914	0.0463	0.0180	72	-2.41	033.00	0
033	Salt as chloride (%)	0619	0.3905	0.0270	0.4914	0.0463	0.0180	72	-2.18	033.01	0
033	Salt as chloride (%)	0505	0.4000	0.0400	0.4914	0.0463	0.0180	72	-1.98	033.03	0
033	Salt as chloride (%)	0656	0.4100	0.0000	0.4914	0.0463	0.0180	72	-1.76	033.03	0
033	Salt as chloride (%)	2181	0.4250	0.0500	0.4914	0.0463	0.0180	72	-1.43	033.99	0
033	Salt as chloride (%)	0897	0.4350	0.0100	0.4914	0.0463	0.0180	72	-1.22	033.00	0
033	Salt as chloride (%)	2009	0.4359	0.0034	0.4914	0.0463	0.0180	72	-1.20	033.00	0
033	Salt as chloride (%)	0885	0.4437	0.0107	0.4914	0.0463	0.0180	72	-1.03	033.00	0
033	Salt as chloride (%)	0045	0.4440	0.0300	0.4914	0.0463	0.0180	72	-1.02	033.00	0
033	Salt as chloride (%)	0504	0.4450	0.0100	0.4914	0.0463	0.0180	72	-1.00	033.00	0
033	Salt as chloride (%)	2196	0.4500	0.0000	0.4914	0.0463	0.0180	72	-0.89	033.00	0
033	Salt as chloride (%)	0878	0.4550	0.0300	0.4914	0.0463	0.0180	72	-0.79	033.00	0
033	Salt as chloride (%)	0884	0.4550	0.0300	0.4914	0.0463	0.0180	72	-0.79	033.00	0
033	Salt as chloride (%)	2193	0.4600	0.0200	0.4914	0.0463	0.0180	72	-0.68	033.00	0
033	Salt as chloride (%)	0123	0.4650	0.0100	0.4914	0.0463	0.0180	72	-0.57	033.00	0
033	Salt as chloride (%)	0511	0.4650	0.0100	0.4914	0.0463	0.0180	72	-0.57	033.00	0
033	Salt as chloride (%)	0893	0.4650	0.0300	0.4914	0.0463	0.0180	72	-0.57	033.00	0
033	Salt as chloride (%)	2076	0.4650	0.0300	0.4914	0.0463	0.0180	72	-0.57	033.00	0
033	Salt as chloride (%)	0083	0.4700	0.0000	0.4914	0.0463	0.0180	72	-0.46	033.01	0
033	Salt as chloride (%)	0650	0.4700	0.0000	0.4914	0.0463	0.0180	72	-0.46	033.01	0
033	Salt as chloride (%)	0309	0.4799	0.0150	0.4914	0.0463	0.0180	72	-0.25	033.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
033	Salt as chloride (%)	0682	0.4800	0.0000	0.4914	0.0463	0.0180	72	-0.25	033.99	0
033	Salt as chloride (%)	0265	0.4850	0.0100	0.4914	0.0463	0.0180	72	-0.14	033.03	0
033	Salt as chloride (%)	0910	0.4900	0.0600	0.4914	0.0463	0.0180	72	-0.03	033.99	0
033	Salt as chloride (%)	0353	0.4950	0.0100	0.4914	0.0463	0.0180	72	0.08	033.00	0
033	Salt as chloride (%)	0098	0.4950	0.0100	0.4914	0.0463	0.0180	72	0.08	033.01	0
033	Salt as chloride (%)	0590	0.4950	0.0100	0.4914	0.0463	0.0180	72	0.08	033.01	0
033	Salt as chloride (%)	0683	0.4950	0.0100	0.4914	0.0463	0.0180	72	0.08	033.01	0
033	Salt as chloride (%)	0948	0.4950	0.0100	0.4914	0.0463	0.0180	72	0.08	033.01	0
033	Salt as chloride (%)	0689	0.4950	0.0100	0.4914	0.0463	0.0180	72	0.08	033.05	0
033	Salt as chloride (%)	2062	0.4951	0.0050	0.4914	0.0463	0.0180	72	0.08	033.00	0
033	Salt as chloride (%)	0003	0.5000	0.0000	0.4914	0.0463	0.0180	72	0.19	033.01	0
033	Salt as chloride (%)	0148	0.5000	0.0000	0.4914	0.0463	0.0180	72	0.19	033.01	0
033	Salt as chloride (%)	0939	0.5000	0.0000	0.4914	0.0463	0.0180	72	0.19	033.01	0
033	Salt as chloride (%)	0190	0.5000	0.0400	0.4914	0.0463	0.0180	72	0.19	033.03	0
033	Salt as chloride (%)	0726	0.5000	0.0000	0.4914	0.0463	0.0180	72	0.19	033.03	0
033	Salt as chloride (%)	0229	0.5025	0.0010	0.4914	0.0463	0.0180	72	0.24	033.01	0
033	Salt as chloride (%)	0366	0.5050	0.0300	0.4914	0.0463	0.0180	72	0.29	033.00	0
033	Salt as chloride (%)	0001	0.5050	0.0100	0.4914	0.0463	0.0180	72	0.29	033.01	0
033	Salt as chloride (%)	0164	0.5050	0.0100	0.4914	0.0463	0.0180	72	0.29	033.01	0
033	Salt as chloride (%)	0354	0.5050	0.0100	0.4914	0.0463	0.0180	72	0.29	033.01	0
033	Salt as chloride (%)	0032	0.5050	0.0500	0.4914	0.0463	0.0180	72	0.29	033.99	0
033	Salt as chloride (%)	0278	0.5100	0.0000	0.4914	0.0463	0.0180	72	0.40	033.01	0
033	Salt as chloride (%)	0425	0.5100	0.0000	0.4914	0.0463	0.0180	72	0.40	033.01	0
033	Salt as chloride (%)	0242	0.5110	0.0000	0.4914	0.0463	0.0180	72	0.42	033.01	0
033	Salt as chloride (%)	0510	0.5110	0.0000	0.4914	0.0463	0.0180	72	0.42	033.01	0
033	Salt as chloride (%)	0881	0.5120	0.0180	0.4914	0.0463	0.0180	72	0.45	033.00	0
033	Salt as chloride (%)	0100	0.5150	0.0300	0.4914	0.0463	0.0180	72	0.51	033.01	0
033	Salt as chloride (%)	0175	0.5150	0.0500	0.4914	0.0463	0.0180	72	0.51	033.01	0
033	Salt as chloride (%)	2109	0.5160	0.0000	0.4914	0.0463	0.0180	72	0.53	033.01	0
033	Salt as chloride (%)	0964	0.5170	0.0020	0.4914	0.0463	0.0180	72	0.55	033.01	0
033	Salt as chloride (%)	0407	0.5185	0.0010	0.4914	0.0463	0.0180	72	0.59	033.01	0
033	Salt as chloride (%)	0004	0.5200	0.0000	0.4914	0.0463	0.0180	72	0.62	033.01	0
033	Salt as chloride (%)	0144	0.5200	0.0400	0.4914	0.0463	0.0180	72	0.62	033.03	0
033	Salt as chloride (%)	0017	0.5285	0.0090	0.4914	0.0463	0.0180	72	0.80	033.99	0
033	Salt as chloride (%)	0539	0.5300	0.0400	0.4914	0.0463	0.0180	72	0.84	033.00	0
033	Salt as chloride (%)	0019	0.5300	0.0000	0.4914	0.0463	0.0180	72	0.84	033.01	0
033	Salt as chloride (%)	0169	0.5350	0.0100	0.4914	0.0463	0.0180	72	0.94	033.00	0
033	Salt as chloride (%)	0629	0.5350	0.0100	0.4914	0.0463	0.0180	72	0.94	033.01	0
033	Salt as chloride (%)	0868	0.5400	0.0000	0.4914	0.0463	0.0180	72	1.05	033.01	0
033	Salt as chloride (%)	0644	0.5460	0.0000	0.4914	0.0463	0.0180	72	1.18	033.01	0
033	Salt as chloride (%)	2190	0.5539	0.0600	0.4914	0.0463	0.0180	72	1.35	033.00	0
033	Salt as chloride (%)	2268	0.5550	0.0300	0.4914	0.0463	0.0180	72	1.38	033.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
033	Salt as chloride (%)	0226	0.5550	0.0100	0.4914	0.0463	0.0180	72	1.38	033.01	0
033	Salt as chloride (%)	0027	0.5710	0.0580	0.4914	0.0463	0.0180	72	1.72	033.99	0
033	Salt as chloride (%)	0874	0.5770	0.0000	0.4914	0.0463	0.0180	72	1.85	033.01	0
033	Salt as chloride (%)	0358	0.5950	0.0700	0.4914	0.0463	0.0180	72	2.24	033.99	0
033	Salt as chloride (%)	2043	0.7750	0.1100	0.4914	0.0463	0.0180	72	6.13	033.05	0
033	Salt as chloride (%)	0693	0.2955	0.3190	0.4914	0.0463	0.0180	72	-4.23	033.00	1
033	Salt as chloride (%)	0536	1.110	0.0400	0.4914	0.0463	0.0180	72	13.37	033.99	2
034	Selenium (mg / kg (ppm))	2234	0.1850	0.0100	0.7704	0.1412	0.0472	20	-4.15	034.53	0
034	Selenium (mg / kg (ppm))	0047	0.5655	0.0870	0.7704	0.1412	0.0472	20	-1.45	034.52	0
034	Selenium (mg / kg (ppm))	0619	0.6350	0.0140	0.7704	0.1412	0.0472	20	-0.96	034.41	0
034	Selenium (mg / kg (ppm))	0910	0.6500	0.0600	0.7704	0.1412	0.0472	20	-0.85	034.52	0
034	Selenium (mg / kg (ppm))	0171	0.6685	0.0510	0.7704	0.1412	0.0472	20	-0.72	034.04	0
034	Selenium (mg / kg (ppm))	0045	0.6775	0.0050	0.7704	0.1412	0.0472	20	-0.66	034.04	0
034	Selenium (mg / kg (ppm))	0968	0.7015	0.0270	0.7704	0.1412	0.0472	20	-0.49	034.43	0
034	Selenium (mg / kg (ppm))	0164	0.7370	0.0120	0.7704	0.1412	0.0472	20	-0.24	034.53	0
034	Selenium (mg / kg (ppm))	0148	0.7400	0.0800	0.7704	0.1412	0.0472	20	-0.22	034.41	0
034	Selenium (mg / kg (ppm))	0941	0.7890	0.0020	0.7704	0.1412	0.0472	20	0.13	034.43	0
034	Selenium (mg / kg (ppm))	0560	0.7895	0.0010	0.7704	0.1412	0.0472	20	0.14	034.53	0
034	Selenium (mg / kg (ppm))	0553	0.7900	0.0020	0.7704	0.1412	0.0472	20	0.14	034.53	0
034	Selenium (mg / kg (ppm))	0038	0.8115	0.0070	0.7704	0.1412	0.0472	20	0.29	034.01	0
034	Selenium (mg / kg (ppm))	0169	0.8150	0.0300	0.7704	0.1412	0.0472	20	0.32	034.04	0
034	Selenium (mg / kg (ppm))	0692	0.8500	0.3000	0.7704	0.1412	0.0472	20	0.56	034.42	0
034	Selenium (mg / kg (ppm))	0918	0.8546	0.0119	0.7704	0.1412	0.0472	20	0.60	034.53	0
034	Selenium (mg / kg (ppm))	0032	0.8550	0.0300	0.7704	0.1412	0.0472	20	0.60	034.53	0
034	Selenium (mg / kg (ppm))	0939	0.9550	0.1900	0.7704	0.1412	0.0472	20	1.31	034.53	0
034	Selenium (mg / kg (ppm))	0407	1.688	0.0150	0.7704	0.1412	0.0472	20	6.49	034.41	0
034	Selenium (mg / kg (ppm))	0190	1.855	0.0100	0.7704	0.1412	0.0472	20	7.68	034.99	0
034	Selenium (mg / kg (ppm))	0407	2.043	0.3650	0.7704	0.1412	0.0472	20	9.01	034.43	1
034	Selenium (mg / kg (ppm))	0011	8.339	0.0000	0.7704	0.1412	0.0472	20	53.59	034.41	2
035	Sodium (%)	2248	0.1600	0.0000	0.2159	0.0176	0.0072	87	-3.19	035.05	0
035	Sodium (%)	0868	0.1770	0.0020	0.2159	0.0176	0.0072	87	-2.22	035.01	0
035	Sodium (%)	0510	0.1800	0.0020	0.2159	0.0176	0.0072	87	-2.05	035.43	0
035	Sodium (%)	0536	0.1815	0.0010	0.2159	0.0176	0.0072	87	-1.96	035.31	0
035	Sodium (%)	2129	0.1830	0.0020	0.2159	0.0176	0.0072	87	-1.88	035.99	0
035	Sodium (%)	0019	0.1900	0.0000	0.2159	0.0176	0.0072	87	-1.48	035.41	0
035	Sodium (%)	0123	0.1900	0.0200	0.2159	0.0176	0.0072	87	-1.48	035.41	0
035	Sodium (%)	0265	0.1900	0.0200	0.2159	0.0176	0.0072	87	-1.48	035.42	0
035	Sodium (%)	0294	0.1900	0.0000	0.2159	0.0176	0.0072	87	-1.48	035.42	0
035	Sodium (%)	0154	0.1933	0.0008	0.2159	0.0176	0.0072	87	-1.29	035.52	0
035	Sodium (%)	0505	0.1950	0.0100	0.2159	0.0176	0.0072	87	-1.19	035.43	0
035	Sodium (%)	0045	0.1970	0.0240	0.2159	0.0176	0.0072	87	-1.08	035.42	0
035	Sodium (%)	2246	0.2000	0.0000	0.2159	0.0176	0.0072	87	-0.91	035.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
035	Sodium (%)	0144	0.2000	0.0200	0.2159	0.0176	0.0072	87	-0.91	035.41	0
035	Sodium (%)	0298	0.2000	0.0200	0.2159	0.0176	0.0072	87	-0.91	035.41	0
035	Sodium (%)	0098	0.2000	0.0000	0.2159	0.0176	0.0072	87	-0.91	035.43	0
035	Sodium (%)	0353	0.2000	0.0000	0.2159	0.0176	0.0072	87	-0.91	035.43	0
035	Sodium (%)	0229	0.2040	0.0100	0.2159	0.0176	0.0072	87	-0.68	035.41	0
035	Sodium (%)	0939	0.2050	0.0100	0.2159	0.0176	0.0072	87	-0.62	035.31	0
035	Sodium (%)	0511	0.2050	0.0100	0.2159	0.0176	0.0072	87	-0.62	035.41	0
035	Sodium (%)	0035	0.2050	0.0100	0.2159	0.0176	0.0072	87	-0.62	035.42	0
035	Sodium (%)	0242	0.2050	0.0100	0.2159	0.0176	0.0072	87	-0.62	035.99	0
035	Sodium (%)	0345	0.2075	0.0050	0.2159	0.0176	0.0072	87	-0.48	035.43	0
035	Sodium (%)	0017	0.2080	0.0040	0.2159	0.0176	0.0072	87	-0.45	035.43	0
035	Sodium (%)	0164	0.2090	0.0040	0.2159	0.0176	0.0072	87	-0.40	035.41	0
035	Sodium (%)	0560	0.2094	0.0002	0.2159	0.0176	0.0072	87	-0.37	035.42	0
035	Sodium (%)	0598	0.2096	0.0047	0.2159	0.0176	0.0072	87	-0.36	035.41	0
035	Sodium (%)	0049	0.2100	0.0000	0.2159	0.0176	0.0072	87	-0.34	035.41	0
035	Sodium (%)	0098	0.2100	0.0000	0.2159	0.0176	0.0072	87	-0.34	035.41	0
035	Sodium (%)	0148	0.2100	0.0000	0.2159	0.0176	0.0072	87	-0.34	035.41	0
035	Sodium (%)	0035	0.2100	0.0000	0.2159	0.0176	0.0072	87	-0.34	035.43	0
035	Sodium (%)	0610	0.2100	0.0000	0.2159	0.0176	0.0072	87	-0.34	035.43	0
035	Sodium (%)	0629	0.2100	0.0000	0.2159	0.0176	0.0072	87	-0.34	035.43	0
035	Sodium (%)	0939	0.2100	0.0000	0.2159	0.0176	0.0072	87	-0.34	035.53	0
035	Sodium (%)	0100	0.2100	0.0000	0.2159	0.0176	0.0072	87	-0.34	035.99	0
035	Sodium (%)	2097	0.2130	0.0120	0.2159	0.0176	0.0072	87	-0.17	035.31	0
035	Sodium (%)	0693	0.2130	0.0000	0.2159	0.0176	0.0072	87	-0.17	035.42	0
035	Sodium (%)	0037	0.2145	0.0070	0.2159	0.0176	0.0072	87	-0.08	035.05	0
035	Sodium (%)	0948	0.2150	0.0020	0.2159	0.0176	0.0072	87	-0.05	035.31	0
035	Sodium (%)	0169	0.2150	0.0100	0.2159	0.0176	0.0072	87	-0.05	035.32	0
035	Sodium (%)	0171	0.2150	0.0100	0.2159	0.0176	0.0072	87	-0.05	035.41	0
035	Sodium (%)	0354	0.2150	0.0100	0.2159	0.0176	0.0072	87	-0.05	035.41	0
035	Sodium (%)	0878	0.2150	0.0100	0.2159	0.0176	0.0072	87	-0.05	035.41	0
035	Sodium (%)	0910	0.2150	0.0100	0.2159	0.0176	0.0072	87	-0.05	035.41	0
035	Sodium (%)	0190	0.2150	0.0100	0.2159	0.0176	0.0072	87	-0.05	035.42	0
035	Sodium (%)	0297	0.2150	0.0100	0.2159	0.0176	0.0072	87	-0.05	035.43	0
035	Sodium (%)	0042	0.2155	0.0050	0.2159	0.0176	0.0072	87	-0.02	035.42	0
035	Sodium (%)	0918	0.2155	0.0030	0.2159	0.0176	0.0072	87	-0.02	035.43	0
035	Sodium (%)	0881	0.2165	0.0010	0.2159	0.0176	0.0072	87	0.03	035.43	0
035	Sodium (%)	0726	0.2168	0.0009	0.2159	0.0176	0.0072	87	0.05	035.42	0
035	Sodium (%)	0142	0.2171	0.0024	0.2159	0.0176	0.0072	87	0.07	035.31	0
035	Sodium (%)	0870	0.2173	0.0042	0.2159	0.0176	0.0072	87	0.08	035.42	0
035	Sodium (%)	0407	0.2173	0.0056	0.2159	0.0176	0.0072	87	0.08	035.43	0
035	Sodium (%)	0520	0.2180	0.0044	0.2159	0.0176	0.0072	87	0.12	035.41	0
035	Sodium (%)	0407	0.2185	0.0046	0.2159	0.0176	0.0072	87	0.15	035.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
035	Sodium (%)	0619	0.2190	0.0200	0.2159	0.0176	0.0072	87	0.17	035.41	0
035	Sodium (%)	2196	0.2200	0.0000	0.2159	0.0176	0.0072	87	0.23	035.31	0
035	Sodium (%)	0083	0.2200	0.0000	0.2159	0.0176	0.0072	87	0.23	035.43	0
035	Sodium (%)	0202	0.2200	0.0040	0.2159	0.0176	0.0072	87	0.23	035.43	0
035	Sodium (%)	0682	0.2200	0.0000	0.2159	0.0176	0.0072	87	0.23	035.43	0
035	Sodium (%)	2192	0.2200	0.0200	0.2159	0.0176	0.0072	87	0.23	035.43	0
035	Sodium (%)	2053	0.2250	0.0300	0.2159	0.0176	0.0072	87	0.52	035.42	0
035	Sodium (%)	2089	0.2250	0.0100	0.2159	0.0176	0.0072	87	0.52	035.43	0
035	Sodium (%)	0590	0.2250	0.0100	0.2159	0.0176	0.0072	87	0.52	035.99	0
035	Sodium (%)	0553	0.2265	0.0410	0.2159	0.0176	0.0072	87	0.60	035.53	0
035	Sodium (%)	0868	0.2300	0.0200	0.2159	0.0176	0.0072	87	0.80	035.31	0
035	Sodium (%)	0425	0.2300	0.0200	0.2159	0.0176	0.0072	87	0.80	035.43	0
035	Sodium (%)	0720	0.2300	0.0000	0.2159	0.0176	0.0072	87	0.80	035.41	0
035	Sodium (%)	0278	0.2300	0.0000	0.2159	0.0176	0.0072	87	0.80	035.42	0
035	Sodium (%)	0089	0.2300	0.0000	0.2159	0.0176	0.0072	87	0.80	035.43	0
035	Sodium (%)	0187	0.2328	0.0006	0.2159	0.0176	0.0072	87	0.96	035.42	0
035	Sodium (%)	0366	0.2335	0.0190	0.2159	0.0176	0.0072	87	1.00	035.42	0
035	Sodium (%)	0629	0.2350	0.0100	0.2159	0.0176	0.0072	87	1.09	035.31	0
035	Sodium (%)	2126	0.2350	0.0100	0.2159	0.0176	0.0072	87	1.09	035.31	0
035	Sodium (%)	0226	0.2350	0.0120	0.2159	0.0176	0.0072	87	1.09	035.41	0
035	Sodium (%)	0358	0.2350	0.0300	0.2159	0.0176	0.0072	87	1.09	035.41	0
035	Sodium (%)	0108	0.2400	0.0000	0.2159	0.0176	0.0072	87	1.37	035.05	0
035	Sodium (%)	0656	0.2400	0.0000	0.2159	0.0176	0.0072	87	1.37	035.31	0
035	Sodium (%)	0683	0.2400	0.0000	0.2159	0.0176	0.0072	87	1.37	035.31	0
035	Sodium (%)	0675	0.2400	0.0000	0.2159	0.0176	0.0072	87	1.37	035.43	0
035	Sodium (%)	2251	0.2450	0.0100	0.2159	0.0176	0.0072	87	1.66	035.05	0
035	Sodium (%)	2022	0.2455	0.0150	0.2159	0.0176	0.0072	87	1.68	035.31	0
035	Sodium (%)	0874	0.2620	0.0060	0.2159	0.0176	0.0072	87	2.62	035.31	0
035	Sodium (%)	2018	0.2650	0.0100	0.2159	0.0176	0.0072	87	2.80	035.31	0
035	Sodium (%)	0689	0.2800	0.0000	0.2159	0.0176	0.0072	87	3.65	035.01	0
035	Sodium (%)	0695	0.2875	0.0210	0.2159	0.0176	0.0072	87	4.08	035.41	0
035	Sodium (%)	0685	0.2900	0.0000	0.2159	0.0176	0.0072	87	4.22	035.42	0
035	Sodium (%)	0692	0.2650	0.0700	0.2159	0.0176	0.0072	87	2.80	035.42	1
035	Sodium (%)	0889	0.4256	0.0452	0.2159	0.0176	0.0072	87	11.95	035.99	2
036	Sulfur (%)	0037	0.1305	0.0130	0.2531	0.0239	0.0084	42	-5.13	036.42	0
036	Sulfur (%)	0229	0.2100	0.0000	0.2531	0.0239	0.0084	42	-1.80	036.42	0
036	Sulfur (%)	0226	0.2150	0.0100	0.2531	0.0239	0.0084	42	-1.59	036.04	0
036	Sulfur (%)	0045	0.2215	0.0010	0.2531	0.0239	0.0084	42	-1.32	036.42	0
036	Sulfur (%)	0229	0.2250	0.0100	0.2531	0.0239	0.0084	42	-1.18	036.04	0
036	Sulfur (%)	0366	0.2250	0.0100	0.2531	0.0239	0.0084	42	-1.18	036.42	0
036	Sulfur (%)	0035	0.2300	0.0400	0.2531	0.0239	0.0084	42	-0.97	036.43	0
036	Sulfur (%)	0870	0.2326	0.0012	0.2531	0.0239	0.0084	42	-0.86	036.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
036	Sulfur (%)	0035	0.2350	0.0100	0.2531	0.0239	0.0084	42	-0.76	036.42	0
036	Sulfur (%)	0265	0.2350	0.0100	0.2531	0.0239	0.0084	42	-0.76	036.42	0
036	Sulfur (%)	0353	0.2350	0.0100	0.2531	0.0239	0.0084	42	-0.76	036.43	0
036	Sulfur (%)	0169	0.2400	0.0000	0.2531	0.0239	0.0084	42	-0.55	036.43	0
036	Sulfur (%)	0910	0.2415	0.0008	0.2531	0.0239	0.0084	42	-0.49	036.52	0
036	Sulfur (%)	0083	0.2450	0.0100	0.2531	0.0239	0.0084	42	-0.34	036.43	0
036	Sulfur (%)	0357	0.2452	0.0003	0.2531	0.0239	0.0084	42	-0.33	036.42	0
036	Sulfur (%)	0708	0.2480	0.0020	0.2531	0.0239	0.0084	42	-0.21	036.42	0
036	Sulfur (%)	2053	0.2500	0.0000	0.2531	0.0239	0.0084	42	-0.13	036.04	0
036	Sulfur (%)	0164	0.2500	0.0000	0.2531	0.0239	0.0084	42	-0.13	036.42	0
036	Sulfur (%)	0032	0.2500	0.0200	0.2531	0.0239	0.0084	42	-0.13	036.43	0
036	Sulfur (%)	0693	0.2510	0.0140	0.2531	0.0239	0.0084	42	-0.09	036.42	0
036	Sulfur (%)	0889	0.2520	0.0005	0.2531	0.0239	0.0084	42	-0.05	036.99	0
036	Sulfur (%)	0407	0.2542	0.0037	0.2531	0.0239	0.0084	42	0.04	036.43	0
036	Sulfur (%)	0629	0.2550	0.0100	0.2531	0.0239	0.0084	42	0.08	036.43	0
036	Sulfur (%)	0202	0.2570	0.0040	0.2531	0.0239	0.0084	42	0.16	036.43	0
036	Sulfur (%)	0726	0.2577	0.0011	0.2531	0.0239	0.0084	42	0.19	036.42	0
036	Sulfur (%)	0098	0.2590	0.0091	0.2531	0.0239	0.0084	42	0.25	036.43	0
036	Sulfur (%)	0345	0.2595	0.0030	0.2531	0.0239	0.0084	42	0.27	036.43	0
036	Sulfur (%)	0148	0.2600	0.0000	0.2531	0.0239	0.0084	42	0.29	036.04	0
036	Sulfur (%)	0692	0.2600	0.0400	0.2531	0.0239	0.0084	42	0.29	036.42	0
036	Sulfur (%)	0297	0.2600	0.0000	0.2531	0.0239	0.0084	42	0.29	036.43	0
036	Sulfur (%)	0187	0.2634	0.0001	0.2531	0.0239	0.0084	42	0.43	036.42	0
036	Sulfur (%)	2129	0.2638	0.0273	0.2531	0.0239	0.0084	42	0.45	036.42	0
036	Sulfur (%)	0553	0.2640	0.0120	0.2531	0.0239	0.0084	42	0.46	036.53	0
036	Sulfur (%)	0171	0.2700	0.0020	0.2531	0.0239	0.0084	42	0.71	036.42	0
036	Sulfur (%)	0190	0.2800	0.0200	0.2531	0.0239	0.0084	42	1.13	036.42	0
036	Sulfur (%)	0242	0.2800	0.0000	0.2531	0.0239	0.0084	42	1.13	036.99	0
036	Sulfur (%)	0560	0.2829	0.0061	0.2531	0.0239	0.0084	42	1.25	036.42	0
036	Sulfur (%)	0918	0.2845	0.0010	0.2531	0.0239	0.0084	42	1.31	036.43	0
036	Sulfur (%)	2089	0.2900	0.0200	0.2531	0.0239	0.0084	42	1.54	036.43	0
036	Sulfur (%)	0278	0.2950	0.0100	0.2531	0.0239	0.0084	42	1.75	036.42	0
036	Sulfur (%)	0510	0.3350	0.0100	0.2531	0.0239	0.0084	42	3.43	036.43	0
036	Sulfur (%)	0294	0.3450	0.0100	0.2531	0.0239	0.0084	42	3.85	036.42	0
036	Sulfur (%)	0038	0.5565	0.0410	0.2531	0.0239	0.0084	42	12.70	036.43	2
037	Zinc (mg / kg (ppm))	0889	111.8	3.622	150.0	13.53	3.942	87	-2.82	037.99	0
037	Zinc (mg / kg (ppm))	0685	123.4	2.050	150.0	13.53	3.942	87	-1.96	037.42	0
037	Zinc (mg / kg (ppm))	0848	125.0	0.5700	150.0	13.53	3.942	87	-1.85	037.41	0
037	Zinc (mg / kg (ppm))	0066	127.1	0.9500	150.0	13.53	3.942	87	-1.69	037.44	0
037	Zinc (mg / kg (ppm))	0047	129.3	15.22	150.0	13.53	3.942	87	-1.53	037.52	0
037	Zinc (mg / kg (ppm))	0878	130.0	2.000	150.0	13.53	3.942	87	-1.48	037.41	0
037	Zinc (mg / kg (ppm))	2188	130.2	18.10	150.0	13.53	3.942	87	-1.47	037.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
037	Zinc (mg / kg (ppm))	2248	131.1	0.0800	150.0	13.53	3.942	87	-1.39	037.31	0
037	Zinc (mg / kg (ppm))	2129	132.3	4.400	150.0	13.53	3.942	87	-1.31	037.99	0
037	Zinc (mg / kg (ppm))	0529	133.6	0.7000	150.0	13.53	3.942	87	-1.21	037.31	0
037	Zinc (mg / kg (ppm))	0345	133.9	1.400	150.0	13.53	3.942	87	-1.19	037.43	0
037	Zinc (mg / kg (ppm))	0520	134.5	5.000	150.0	13.53	3.942	87	-1.14	037.41	0
037	Zinc (mg / kg (ppm))	0870	134.6	8.750	150.0	13.53	3.942	87	-1.13	037.42	0
037	Zinc (mg / kg (ppm))	0017	135.0	5.223	150.0	13.53	3.942	87	-1.10	037.43	0
037	Zinc (mg / kg (ppm))	0353	135.5	7.000	150.0	13.53	3.942	87	-1.07	037.43	0
037	Zinc (mg / kg (ppm))	0366	136.0	2.000	150.0	13.53	3.942	87	-1.03	037.42	0
037	Zinc (mg / kg (ppm))	0164	137.0	0.0000	150.0	13.53	3.942	87	-0.96	037.41	0
037	Zinc (mg / kg (ppm))	2234	139.5	4.280	150.0	13.53	3.942	87	-0.77	037.43	0
037	Zinc (mg / kg (ppm))	0941	140.0	0.0000	150.0	13.53	3.942	87	-0.74	037.43	0
037	Zinc (mg / kg (ppm))	0045	140.5	1.000	150.0	13.53	3.942	87	-0.70	037.42	0
037	Zinc (mg / kg (ppm))	0026	141.5	5.295	150.0	13.53	3.942	87	-0.63	037.42	0
037	Zinc (mg / kg (ppm))	0242	141.5	5.000	150.0	13.53	3.942	87	-0.63	037.99	0
037	Zinc (mg / kg (ppm))	0689	141.7	12.60	150.0	13.53	3.942	87	-0.61	037.31	0
037	Zinc (mg / kg (ppm))	0910	142.0	6.000	150.0	13.53	3.942	87	-0.59	037.41	0
037	Zinc (mg / kg (ppm))	2022	142.5	3.000	150.0	13.53	3.942	87	-0.55	037.31	0
037	Zinc (mg / kg (ppm))	0098	143.7	0.3000	150.0	13.53	3.942	87	-0.47	037.41	0
037	Zinc (mg / kg (ppm))	0505	143.8	5.100	150.0	13.53	3.942	87	-0.46	037.43	0
037	Zinc (mg / kg (ppm))	0171	144.0	0.0000	150.0	13.53	3.942	87	-0.44	037.41	0
037	Zinc (mg / kg (ppm))	0169	144.0	0.0000	150.0	13.53	3.942	87	-0.44	037.43	0
037	Zinc (mg / kg (ppm))	0354	144.1	0.5700	150.0	13.53	3.942	87	-0.44	037.41	0
037	Zinc (mg / kg (ppm))	0425	145.2	0.1800	150.0	13.53	3.942	87	-0.35	037.43	0
037	Zinc (mg / kg (ppm))	0278	145.9	5.200	150.0	13.53	3.942	87	-0.30	037.42	0
037	Zinc (mg / kg (ppm))	2053	146.5	7.000	150.0	13.53	3.942	87	-0.26	037.42	0
037	Zinc (mg / kg (ppm))	0675	146.9	3.070	150.0	13.53	3.942	87	-0.23	037.43	0
037	Zinc (mg / kg (ppm))	0693	147.4	0.7330	150.0	13.53	3.942	87	-0.19	037.42	0
037	Zinc (mg / kg (ppm))	0629	147.5	1.000	150.0	13.53	3.942	87	-0.18	037.43	0
037	Zinc (mg / kg (ppm))	0049	147.7	5.930	150.0	13.53	3.942	87	-0.17	037.41	0
037	Zinc (mg / kg (ppm))	0074	148.0	10.00	150.0	13.53	3.942	87	-0.15	037.41	0
037	Zinc (mg / kg (ppm))	0019	148.3	2.500	150.0	13.53	3.942	87	-0.13	037.41	0
037	Zinc (mg / kg (ppm))	0939	148.5	0.9100	150.0	13.53	3.942	87	-0.11	037.53	0
037	Zinc (mg / kg (ppm))	0918	149.1	2.910	150.0	13.53	3.942	87	-0.07	037.43	0
037	Zinc (mg / kg (ppm))	0619	149.5	3.000	150.0	13.53	3.942	87	-0.04	037.41	0
037	Zinc (mg / kg (ppm))	0042	149.5	13.00	150.0	13.53	3.942	87	-0.04	037.42	0
037	Zinc (mg / kg (ppm))	0868	150.0	4.000	150.0	13.53	3.942	87	0.00	037.31	0
037	Zinc (mg / kg (ppm))	0510	151.0	0.0000	150.0	13.53	3.942	87	0.08	037.43	0
037	Zinc (mg / kg (ppm))	0202	151.1	7.144	150.0	13.53	3.942	87	0.08	037.43	0
037	Zinc (mg / kg (ppm))	0265	151.5	17.00	150.0	13.53	3.942	87	0.11	037.42	0
037	Zinc (mg / kg (ppm))	0720	151.9	0.2800	150.0	13.53	3.942	87	0.14	037.41	0
037	Zinc (mg / kg (ppm))	0629	152.0	0.0000	150.0	13.53	3.942	87	0.15	037.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
037	Zinc (mg / kg (ppm))	0968	153.0	2.000	150.0	13.53	3.942	87	0.22	037.43	0
037	Zinc (mg / kg (ppm))	0656	153.2	11.61	150.0	13.53	3.942	87	0.24	037.31	0
037	Zinc (mg / kg (ppm))	2196	153.5	0.0000	150.0	13.53	3.942	87	0.26	037.31	0
037	Zinc (mg / kg (ppm))	0190	153.8	1.120	150.0	13.53	3.942	87	0.28	037.42	0
037	Zinc (mg / kg (ppm))	0560	154.3	0.2000	150.0	13.53	3.942	87	0.32	037.42	0
037	Zinc (mg / kg (ppm))	0083	154.5	5.000	150.0	13.53	3.942	87	0.33	037.43	0
037	Zinc (mg / kg (ppm))	2089	154.7	9.840	150.0	13.53	3.942	87	0.34	037.43	0
037	Zinc (mg / kg (ppm))	0536	154.9	6.000	150.0	13.53	3.942	87	0.36	037.31	0
037	Zinc (mg / kg (ppm))	0682	155.1	0.0000	150.0	13.53	3.942	87	0.38	037.43	0
037	Zinc (mg / kg (ppm))	0358	155.4	16.65	150.0	13.53	3.942	87	0.40	037.41	0
037	Zinc (mg / kg (ppm))	0187	156.2	0.9000	150.0	13.53	3.942	87	0.46	037.42	0
037	Zinc (mg / kg (ppm))	0038	156.5	1.000	150.0	13.53	3.942	87	0.48	037.43	0
037	Zinc (mg / kg (ppm))	0297	156.5	1.000	150.0	13.53	3.942	87	0.48	037.43	0
037	Zinc (mg / kg (ppm))	0598	156.6	0.2000	150.0	13.53	3.942	87	0.49	037.41	0
037	Zinc (mg / kg (ppm))	0692	157.7	3.100	150.0	13.53	3.942	87	0.57	037.42	0
037	Zinc (mg / kg (ppm))	0939	158.0	1.000	150.0	13.53	3.942	87	0.59	037.31	0
037	Zinc (mg / kg (ppm))	0038	158.0	2.000	150.0	13.53	3.942	87	0.59	037.32	0
037	Zinc (mg / kg (ppm))	0407	159.5	2.380	150.0	13.53	3.942	87	0.70	037.41	0
037	Zinc (mg / kg (ppm))	0357	159.9	0.0610	150.0	13.53	3.942	87	0.73	037.42	0
037	Zinc (mg / kg (ppm))	0726	160.0	9.300	150.0	13.53	3.942	87	0.74	037.42	0
037	Zinc (mg / kg (ppm))	0407	161.3	3.528	150.0	13.53	3.942	87	0.84	037.43	0
037	Zinc (mg / kg (ppm))	0226	161.5	3.000	150.0	13.53	3.942	87	0.85	037.41	0
037	Zinc (mg / kg (ppm))	0100	163.0	4.000	150.0	13.53	3.942	87	0.96	037.99	0
037	Zinc (mg / kg (ppm))	0148	163.8	2.970	150.0	13.53	3.942	87	1.02	037.41	0
037	Zinc (mg / kg (ppm))	0098	164.0	5.300	150.0	13.53	3.942	87	1.03	037.43	0
037	Zinc (mg / kg (ppm))	0874	164.5	1.000	150.0	13.53	3.942	87	1.07	037.31	0
037	Zinc (mg / kg (ppm))	0032	164.5	9.000	150.0	13.53	3.942	87	1.07	037.43	0
037	Zinc (mg / kg (ppm))	0294	165.3	3.100	150.0	13.53	3.942	87	1.13	037.42	0
037	Zinc (mg / kg (ppm))	0001	165.3	4.400	150.0	13.53	3.942	87	1.13	037.31	0
037	Zinc (mg / kg (ppm))	0035	165.5	1.000	150.0	13.53	3.942	87	1.15	037.42	0
037	Zinc (mg / kg (ppm))	0511	169.0	4.000	150.0	13.53	3.942	87	1.41	037.41	0
037	Zinc (mg / kg (ppm))	0035	170.0	2.000	150.0	13.53	3.942	87	1.48	037.43	0
037	Zinc (mg / kg (ppm))	0948	171.9	3.089	150.0	13.53	3.942	87	1.62	037.33	0
037	Zinc (mg / kg (ppm))	0590	176.0	10.00	150.0	13.53	3.942	87	1.92	037.99	0
037	Zinc (mg / kg (ppm))	0229	176.7	3.800	150.0	13.53	3.942	87	1.97	037.41	0
037	Zinc (mg / kg (ppm))	0553	179.0	4.000	150.0	13.53	3.942	87	2.14	037.53	0
037	Zinc (mg / kg (ppm))	0037	180.7	1.300	150.0	13.53	3.942	87	2.27	037.42	0
037	Zinc (mg / kg (ppm))	0011	486.5	0.0000	150.0	13.53	3.942	87	24.86	037.41	0
037	Zinc (mg / kg (ppm))	2246	150.0	100.0	150.0	13.53	3.942	87	0.00	037.33	1
037	Zinc (mg / kg (ppm))	0003	196.0	66.00	150.0	13.53	3.942	87	3.40	037.41	1
037	Zinc (mg / kg (ppm))	2178	2,985	4.000	150.0	13.53	3.942	87	209.47	037.33	2
038	Molybdenum (mg / kg (ppm))	0297	0.8300	0.1400	1.203	0.1675	0.1626	18	-2.23	038.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
038	Molybdenum (mg / kg (ppm))	0353	0.8664	0.2021	1.203	0.1675	0.1626	18	-2.01	038.43	0
038	Molybdenum (mg / kg (ppm))	0910	0.9100	0.2000	1.203	0.1675	0.1626	18	-1.75	038.52	0
038	Molybdenum (mg / kg (ppm))	0171	1.080	0.0220	1.203	0.1675	0.1626	18	-0.74	038.41	0
038	Molybdenum (mg / kg (ppm))	0345	1.125	0.0300	1.203	0.1675	0.1626	18	-0.47	038.43	0
038	Molybdenum (mg / kg (ppm))	0169	1.150	0.1000	1.203	0.1675	0.1626	18	-0.32	038.43	0
038	Molybdenum (mg / kg (ppm))	0693	1.178	0.0850	1.203	0.1675	0.1626	18	-0.15	038.42	0
038	Molybdenum (mg / kg (ppm))	0510	1.200	0.0000	1.203	0.1675	0.1626	18	-0.02	038.43	0
038	Molybdenum (mg / kg (ppm))	0407	1.210	0.0370	1.203	0.1675	0.1626	18	0.04	038.41	0
038	Molybdenum (mg / kg (ppm))	0226	1.230	0.1200	1.203	0.1675	0.1626	18	0.16	038.41	0
038	Molybdenum (mg / kg (ppm))	0038	1.248	0.3850	1.203	0.1675	0.1626	18	0.26	038.43	0
038	Molybdenum (mg / kg (ppm))	0164	1.280	0.1200	1.203	0.1675	0.1626	18	0.46	038.53	0
038	Molybdenum (mg / kg (ppm))	0278	1.290	0.1400	1.203	0.1675	0.1626	18	0.52	038.42	0
038	Molybdenum (mg / kg (ppm))	0918	1.297	0.0084	1.203	0.1675	0.1626	18	0.56	038.53	0
038	Molybdenum (mg / kg (ppm))	0407	1.301	0.1877	1.203	0.1675	0.1626	18	0.58	038.43	0
038	Molybdenum (mg / kg (ppm))	0083	1.400	0.0000	1.203	0.1675	0.1626	18	1.17	038.43	0
038	Molybdenum (mg / kg (ppm))	0553	1.400	0.0400	1.203	0.1675	0.1626	18	1.17	038.53	0
038	Molybdenum (mg / kg (ppm))	0265	1.415	1.110	1.203	0.1675	0.1626	18	1.26	038.42	0
038	Molybdenum (mg / kg (ppm))	0045	4.300	2.460	1.203	0.1675	0.1626	18	18.49	038.42	2
040	Barium (mg / kg (ppm))	0918	11.79	0.3820				3		040.53	0
040	Barium (mg / kg (ppm))	0560	12.55	0.0900				3		040.42	0
040	Barium (mg / kg (ppm))	0032	19.20	2.000				3		040.43	0
041	Vanadium (mg / kg (ppm))	0553	0.9110	0.0940				1		041.53	0
042	Chloride (%)	0948	0.3010	0.0060				2		042.00	0
042	Chloride (%)	0297	0.3050	0.0300				2		042.99	0
102	Niacin (mg / kg (ppm))	0939	83.85	2.300				2		102.02	0
102	Niacin (mg / kg (ppm))	0227	136.0	0.0000				2		102.01	0
103	Pantothenic Acid (mg / kg (ppm))	0227	20.45	1.100				2		103.01	0
103	Pantothenic Acid (mg / kg (ppm))	0939	22.45	1.900				2		103.02	0
104	Riboflavin (mg / kg (ppm))	0910	13.99	0.8100	19.40	4.010	0.7650	4		104.03	0
104	Riboflavin (mg / kg (ppm))	0227	19.20	1.600	19.40	4.010	0.7650	4		104.00	0
104	Riboflavin (mg / kg (ppm))	0939	20.98	0.3500	19.40	4.010	0.7650	4		104.03	0
104	Riboflavin (mg / kg (ppm))	0171	23.45	0.3000	19.40	4.010	0.7650	4		104.00	0
105	Thiamine (mg / kg (ppm))	0939	4.480	0.7400				2		105.00	0
105	Thiamine (mg / kg (ppm))	0910	5.310	0.2400				2		105.00	0
106	Vitamin A (KU / kg)	0169	3.000	0.0000	14.65	3.728	1.456	18	-3.12	106.02	0
106	Vitamin A (KU / kg)	0038	9.616	2.186	14.65	3.728	1.456	18	-1.35	106.02	0
106	Vitamin A (KU / kg)	0227	11.14	3.320	14.65	3.728	1.456	18	-0.94	106.02	0
106	Vitamin A (KU / kg)	0905	11.34	0.1130	14.65	3.728	1.456	18	-0.89	106.02	0
106	Vitamin A (KU / kg)	0941	12.01	3.507	14.65	3.728	1.456	18	-0.71	106.02	0
106	Vitamin A (KU / kg)	0004	13.17	0.3700	14.65	3.728	1.456	18	-0.40	106.02	0
106	Vitamin A (KU / kg)	0689	14.05	0.7000	14.65	3.728	1.456	18	-0.16	106.02	0
106	Vitamin A (KU / kg)	0910	14.05	1.700	14.65	3.728	1.456	18	-0.16	106.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
106	Vitamin A (KU / kg)	0656	14.20	1.480	14.65	3.728	1.456	18	-0.12	106.02	0
106	Vitamin A (KU / kg)	0610	14.25	0.5000	14.65	3.728	1.456	18	-0.11	106.02	0
106	Vitamin A (KU / kg)	2053	15.10	2.600	14.65	3.728	1.456	18	0.12	106.02	0
106	Vitamin A (KU / kg)	2103	15.67	1.740	14.65	3.728	1.456	18	0.27	106.02	0
106	Vitamin A (KU / kg)	0676	16.55	0.5000	14.65	3.728	1.456	18	0.51	106.02	0
106	Vitamin A (KU / kg)	0148	17.41	0.2100	14.65	3.728	1.456	18	0.74	106.02	0
106	Vitamin A (KU / kg)	0010	17.67	0.8600	14.65	3.728	1.456	18	0.81	106.02	0
106	Vitamin A (KU / kg)	0904	18.20	3.290	14.65	3.728	1.456	18	0.95	106.99	0
106	Vitamin A (KU / kg)	0017	19.96	1.240	14.65	3.728	1.456	18	1.42	106.02	0
106	Vitamin A (KU / kg)	0098	20.55	1.900	14.65	3.728	1.456	18	1.58	106.01	0
106	Vitamin A (KU / kg)	0675	41.27	10.32	14.65	3.728	1.456	18	7.14	106.02	1
107	Vitamin B12 (µg / kg (ppb))	0227	57.85	47.10				1		107.00	0
108	Vitamin D3 (KU / kg)	0676	3.515	0.0100				3		108.02	0
108	Vitamin D3 (KU / kg)	0227	4.830	0.0000				3		108.02	0
108	Vitamin D3 (KU / kg)	0675	9.775	0.1100				3		108.02	0
108	Vitamin D3 (KU / kg)	0910	4.570	0.3200				3		108.02	1
108	Vitamin D3 (KU / kg)	0169	0.0000	0.0000				3		108.01	4
108	Vitamin D3 (KU / kg)	0610	< 5					3		108.02	5
109	Vitamin E (IU / kg)	0619	0.0505	0.0010	39.73	13.59	3.162	16	-2.92	109.02	0
109	Vitamin E (IU / kg)	0227	4.835	0.2100	39.73	13.59	3.162	16	-2.57	109.02	0
109	Vitamin E (IU / kg)	0169	7.500	1.000	39.73	13.59	3.162	16	-2.37	109.02	0
109	Vitamin E (IU / kg)	0905	35.48	0.5700	39.73	13.59	3.162	16	-0.31	109.02	0
109	Vitamin E (IU / kg)	0904	36.32	4.580	39.73	13.59	3.162	16	-0.25	109.99	0
109	Vitamin E (IU / kg)	2103	37.50	3.000	39.73	13.59	3.162	16	-0.16	109.02	0
109	Vitamin E (IU / kg)	0098	37.55	8.100	39.73	13.59	3.162	16	-0.16	109.02	0
109	Vitamin E (IU / kg)	0941	40.49	1.570	39.73	13.59	3.162	16	0.06	109.02	0
109	Vitamin E (IU / kg)	2053	41.00	8.000	39.73	13.59	3.162	16	0.09	109.02	0
109	Vitamin E (IU / kg)	0610	45.00	2.000	39.73	13.59	3.162	16	0.39	109.02	0
109	Vitamin E (IU / kg)	0171	47.50	3.000	39.73	13.59	3.162	16	0.57	109.99	0
109	Vitamin E (IU / kg)	0910	47.95	0.1000	39.73	13.59	3.162	16	0.61	109.02	0
109	Vitamin E (IU / kg)	2055	47.95	7.700	39.73	13.59	3.162	16	0.61	109.02	0
109	Vitamin E (IU / kg)	0676	50.10	0.8000	39.73	13.59	3.162	16	0.76	109.02	0
109	Vitamin E (IU / kg)	0148	50.66	4.150	39.73	13.59	3.162	16	0.80	109.02	0
109	Vitamin E (IU / kg)	0675	79.74	5.810	39.73	13.59	3.162	16	2.94	109.02	0
112	Pyridoxine (µg / g)	0227	8.640	0.0400				1		112.01	0
113	Folic Acid (mg / kg (ppm))	0939	3.210	0.0200				2		113.02	0
113	Folic Acid (mg / kg (ppm))	0227	4.750	0.1400				2		113.01	0
114	Biotin (mg / kg (ppm))	0227	0.5545	0.2110				1		114.01	0
118	Peroxide value (meq/kg)	2081	5.305	0.3300				1		118.99	0
120	Alanine (%)	0353	1.055	0.0100	1.182	0.0425	0.0147	33	-2.98	120.00	0
120	Alanine (%)	0918	1.081	0.0127	1.182	0.0425	0.0147	33	-2.36	120.00	0
120	Alanine (%)	0407	1.094	0.0110	1.182	0.0425	0.0147	33	-2.08	120.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
120	Alanine (%)	0619	1.120	0.0000	1.182	0.0425	0.0147	33	-1.45	120.00	0
120	Alanine (%)	0889	1.121	0.0137	1.182	0.0425	0.0147	33	-1.43	120.99	0
120	Alanine (%)	0610	1.150	0.0200	1.182	0.0425	0.0147	33	-0.75	120.05	0
120	Alanine (%)	0171	1.154	0.0240	1.182	0.0425	0.0147	33	-0.65	120.00	0
120	Alanine (%)	0878	1.155	0.0020	1.182	0.0425	0.0147	33	-0.63	120.00	0
120	Alanine (%)	0652	1.165	0.0100	1.182	0.0425	0.0147	33	-0.39	120.00	0
120	Alanine (%)	0148	1.170	0.0400	1.182	0.0425	0.0147	33	-0.28	120.05	0
120	Alanine (%)	0676	1.170	0.0200	1.182	0.0425	0.0147	33	-0.28	120.05	0
120	Alanine (%)	2059	1.171	0.0010	1.182	0.0425	0.0147	33	-0.26	120.00	0
120	Alanine (%)	0098	1.173	0.0040	1.182	0.0425	0.0147	33	-0.20	120.02	0
120	Alanine (%)	0644	1.174	0.0190	1.182	0.0425	0.0147	33	-0.19	120.00	0
120	Alanine (%)	0227	1.175	0.0100	1.182	0.0425	0.0147	33	-0.16	120.00	0
120	Alanine (%)	2196	1.176	0.0000	1.182	0.0425	0.0147	33	-0.13	120.05	0
120	Alanine (%)	0504	1.180	0.0200	1.182	0.0425	0.0147	33	-0.04	120.00	0
120	Alanine (%)	0872	1.182	0.0170	1.182	0.0425	0.0147	33	0.00	120.00	0
120	Alanine (%)	0571	1.184	0.0020	1.182	0.0425	0.0147	33	0.05	120.00	0
120	Alanine (%)	0868	1.186	0.0040	1.182	0.0425	0.0147	33	0.10	120.00	0
120	Alanine (%)	0626	1.189	0.0600	1.182	0.0425	0.0147	33	0.17	120.05	0
120	Alanine (%)	0870	1.190	0.0268	1.182	0.0425	0.0147	33	0.21	120.00	0
120	Alanine (%)	2246	1.205	0.0100	1.182	0.0425	0.0147	33	0.55	120.05	0
120	Alanine (%)	0859	1.207	0.0017	1.182	0.0425	0.0147	33	0.59	120.00	0
120	Alanine (%)	0941	1.208	0.0000	1.182	0.0425	0.0147	33	0.62	120.00	0
120	Alanine (%)	0354	1.208	0.0060	1.182	0.0425	0.0147	33	0.62	120.00	0
120	Alanine (%)	0242	1.215	0.0300	1.182	0.0425	0.0147	33	0.78	120.05	0
120	Alanine (%)	0968	1.225	0.0050	1.182	0.0425	0.0147	33	1.01	120.00	0
120	Alanine (%)	0675	1.230	0.0200	1.182	0.0425	0.0147	33	1.14	120.00	0
120	Alanine (%)	0682	1.230	0.0000	1.182	0.0425	0.0147	33	1.14	120.00	0
120	Alanine (%)	0684	1.240	0.0200	1.182	0.0425	0.0147	33	1.37	120.00	0
120	Alanine (%)	0226	1.298	0.0666	1.182	0.0425	0.0147	33	2.74	120.00	0
120	Alanine (%)	2188	1.320	0.0000	1.182	0.0425	0.0147	33	3.26	120.05	0
120	Alanine (%)	0910	1.125	0.0900	1.182	0.0425	0.0147	33	-1.33	120.00	1
121	Arginine (%)	0407	1.166	0.0070	1.447	0.0646	0.0266	33	-4.36	121.05	0
121	Arginine (%)	0889	1.300	0.0391	1.447	0.0646	0.0266	33	-2.28	121.99	0
121	Arginine (%)	0353	1.335	0.0100	1.447	0.0646	0.0266	33	-1.74	121.00	0
121	Arginine (%)	0675	1.360	0.0200	1.447	0.0646	0.0266	33	-1.35	121.00	0
121	Arginine (%)	0918	1.369	0.0264	1.447	0.0646	0.0266	33	-1.21	121.00	0
121	Arginine (%)	0148	1.370	0.0800	1.447	0.0646	0.0266	33	-1.20	121.05	0
121	Arginine (%)	0171	1.391	0.0270	1.447	0.0646	0.0266	33	-0.88	121.00	0
121	Arginine (%)	0870	1.396	0.0314	1.447	0.0646	0.0266	33	-0.79	121.00	0
121	Arginine (%)	0227	1.400	0.0600	1.447	0.0646	0.0266	33	-0.73	121.00	0
121	Arginine (%)	0619	1.425	0.0100	1.447	0.0646	0.0266	33	-0.35	121.00	0
121	Arginine (%)	0626	1.427	0.0940	1.447	0.0646	0.0266	33	-0.32	121.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
121	Arginine (%)	0910	1.430	0.0600	1.447	0.0646	0.0266	33	-0.27	121.00	0
121	Arginine (%)	0098	1.431	0.0250	1.447	0.0646	0.0266	33	-0.26	121.02	0
121	Arginine (%)	0504	1.450	0.0200	1.447	0.0646	0.0266	33	0.04	121.00	0
121	Arginine (%)	0868	1.453	0.0110	1.447	0.0646	0.0266	33	0.08	121.00	0
121	Arginine (%)	0354	1.454	0.0130	1.447	0.0646	0.0266	33	0.10	121.00	0
121	Arginine (%)	0968	1.458	0.0150	1.447	0.0646	0.0266	33	0.16	121.00	0
121	Arginine (%)	0878	1.463	0.0100	1.447	0.0646	0.0266	33	0.24	121.00	0
121	Arginine (%)	0859	1.467	0.0075	1.447	0.0646	0.0266	33	0.30	121.00	0
121	Arginine (%)	0644	1.468	0.0130	1.447	0.0646	0.0266	33	0.31	121.00	0
121	Arginine (%)	0872	1.469	0.0020	1.447	0.0646	0.0266	33	0.34	121.00	0
121	Arginine (%)	2059	1.473	0.0030	1.447	0.0646	0.0266	33	0.39	121.00	0
121	Arginine (%)	0682	1.475	0.0000	1.447	0.0646	0.0266	33	0.43	121.00	0
121	Arginine (%)	0571	1.477	0.0090	1.447	0.0646	0.0266	33	0.45	121.00	0
121	Arginine (%)	0652	1.480	0.0000	1.447	0.0646	0.0266	33	0.51	121.00	0
121	Arginine (%)	0610	1.485	0.0100	1.447	0.0646	0.0266	33	0.58	121.05	0
121	Arginine (%)	0684	1.490	0.0000	1.447	0.0646	0.0266	33	0.66	121.00	0
121	Arginine (%)	0242	1.505	0.0700	1.447	0.0646	0.0266	33	0.89	121.05	0
121	Arginine (%)	0941	1.516	0.0220	1.447	0.0646	0.0266	33	1.06	121.00	0
121	Arginine (%)	2188	1.520	0.0000	1.447	0.0646	0.0266	33	1.12	121.05	0
121	Arginine (%)	2196	1.525	0.0000	1.447	0.0646	0.0266	33	1.20	121.05	0
121	Arginine (%)	0676	1.595	0.1100	1.447	0.0646	0.0266	33	2.29	121.05	0
121	Arginine (%)	0226	1.610	0.0736	1.447	0.0646	0.0266	33	2.52	121.00	0
122	Aspartic (%)	0610	1.485	0.0100	2.011	0.0720	0.0298	32	-7.31	122.05	0
122	Aspartic (%)	0889	1.704	0.0218	2.011	0.0720	0.0298	32	-4.26	122.99	0
122	Aspartic (%)	0353	1.845	0.0100	2.011	0.0720	0.0298	32	-2.30	122.00	0
122	Aspartic (%)	0407	1.872	0.0110	2.011	0.0720	0.0298	32	-1.94	122.05	0
122	Aspartic (%)	0675	1.900	0.0200	2.011	0.0720	0.0298	32	-1.54	122.00	0
122	Aspartic (%)	2246	1.925	0.0500	2.011	0.0720	0.0298	32	-1.19	122.05	0
122	Aspartic (%)	0626	1.954	0.1310	2.011	0.0720	0.0298	32	-0.80	122.05	0
122	Aspartic (%)	0227	1.990	0.0400	2.011	0.0720	0.0298	32	-0.29	122.00	0
122	Aspartic (%)	0652	1.990	0.0200	2.011	0.0720	0.0298	32	-0.29	122.00	0
122	Aspartic (%)	2196	1.990	0.0000	2.011	0.0720	0.0298	32	-0.29	122.05	0
122	Aspartic (%)	0644	1.997	0.0280	2.011	0.0720	0.0298	32	-0.19	122.00	0
122	Aspartic (%)	0354	2.000	0.0400	2.011	0.0720	0.0298	32	-0.15	122.00	0
122	Aspartic (%)	0910	2.005	0.0700	2.011	0.0720	0.0298	32	-0.08	122.00	0
122	Aspartic (%)	2059	2.006	0.0050	2.011	0.0720	0.0298	32	-0.08	122.00	0
122	Aspartic (%)	0098	2.008	0.0430	2.011	0.0720	0.0298	32	-0.05	122.02	0
122	Aspartic (%)	0870	2.020	0.0228	2.011	0.0720	0.0298	32	0.13	122.00	0
122	Aspartic (%)	0878	2.024	0.0100	2.011	0.0720	0.0298	32	0.18	122.00	0
122	Aspartic (%)	0968	2.028	0.0100	2.011	0.0720	0.0298	32	0.24	122.00	0
122	Aspartic (%)	0571	2.030	0.0250	2.011	0.0720	0.0298	32	0.26	122.00	0
122	Aspartic (%)	0504	2.030	0.0200	2.011	0.0720	0.0298	32	0.27	122.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
122	Aspartic (%)	0682	2.031	0.0000	2.011	0.0720	0.0298	32	0.28	122.00	0
122	Aspartic (%)	0872	2.040	0.0390	2.011	0.0720	0.0298	32	0.40	122.00	0
122	Aspartic (%)	0941	2.045	0.0150	2.011	0.0720	0.0298	32	0.47	122.00	0
122	Aspartic (%)	0859	2.047	0.0105	2.011	0.0720	0.0298	32	0.51	122.00	0
122	Aspartic (%)	0619	2.050	0.0200	2.011	0.0720	0.0298	32	0.54	122.00	0
122	Aspartic (%)	0868	2.055	0.0070	2.011	0.0720	0.0298	32	0.61	122.00	0
122	Aspartic (%)	0918	2.055	0.0336	2.011	0.0720	0.0298	32	0.61	122.00	0
122	Aspartic (%)	0676	2.080	0.0400	2.011	0.0720	0.0298	32	0.96	122.05	0
122	Aspartic (%)	0684	2.100	0.0400	2.011	0.0720	0.0298	32	1.24	122.00	0
122	Aspartic (%)	0148	2.100	0.0400	2.011	0.0720	0.0298	32	1.24	122.05	0
122	Aspartic (%)	0242	2.140	0.0400	2.011	0.0720	0.0298	32	1.79	122.05	0
122	Aspartic (%)	0226	2.299	0.0805	2.011	0.0720	0.0298	32	4.00	122.00	0
122	Aspartic (%)	2188	2.150	0.2400	2.011	0.0720	0.0298	32	1.93	122.05	1
122	Aspartic (%)	0171	2.403	0.2160	2.011	0.0720	0.0298	32	5.45	122.00	1
124	Cysteine/Cystine (%)	2246	0.1300	0.0200	0.3199	0.0329	0.0109	32	-5.77	124.05	0
124	Cysteine/Cystine (%)	0870	0.2752	0.0043	0.3199	0.0329	0.0109	32	-1.36	124.00	0
124	Cysteine/Cystine (%)	0918	0.2800	0.0200	0.3199	0.0329	0.0109	32	-1.21	124.00	0
124	Cysteine/Cystine (%)	0227	0.2850	0.0100	0.3199	0.0329	0.0109	32	-1.06	124.00	0
124	Cysteine/Cystine (%)	2196	0.2870	0.0000	0.3199	0.0329	0.0109	32	-1.00	124.05	0
124	Cysteine/Cystine (%)	0226	0.2937	0.0117	0.3199	0.0329	0.0109	32	-0.80	124.00	0
124	Cysteine/Cystine (%)	0910	0.2950	0.0300	0.3199	0.0329	0.0109	32	-0.76	124.00	0
124	Cysteine/Cystine (%)	0171	0.2960	0.0280	0.3199	0.0329	0.0109	32	-0.73	124.00	0
124	Cysteine/Cystine (%)	0644	0.3015	0.0030	0.3199	0.0329	0.0109	32	-0.56	124.00	0
124	Cysteine/Cystine (%)	0652	0.3050	0.0300	0.3199	0.0329	0.0109	32	-0.45	124.00	0
124	Cysteine/Cystine (%)	0676	0.3050	0.0100	0.3199	0.0329	0.0109	32	-0.45	124.05	0
124	Cysteine/Cystine (%)	0872	0.3055	0.0090	0.3199	0.0329	0.0109	32	-0.44	124.00	0
124	Cysteine/Cystine (%)	0878	0.3055	0.0050	0.3199	0.0329	0.0109	32	-0.44	124.00	0
124	Cysteine/Cystine (%)	0859	0.3086	0.0018	0.3199	0.0329	0.0109	32	-0.34	124.00	0
124	Cysteine/Cystine (%)	0571	0.3130	0.0000	0.3199	0.0329	0.0109	32	-0.21	124.00	0
124	Cysteine/Cystine (%)	0941	0.3170	0.0000	0.3199	0.0329	0.0109	32	-0.09	124.00	0
124	Cysteine/Cystine (%)	0354	0.3185	0.0070	0.3199	0.0329	0.0109	32	-0.04	124.00	0
124	Cysteine/Cystine (%)	0868	0.3185	0.0050	0.3199	0.0329	0.0109	32	-0.04	124.00	0
124	Cysteine/Cystine (%)	2059	0.3195	0.0010	0.3199	0.0329	0.0109	32	-0.01	124.00	0
124	Cysteine/Cystine (%)	0353	0.3250	0.0100	0.3199	0.0329	0.0109	32	0.16	124.00	0
124	Cysteine/Cystine (%)	0098	0.3255	0.0110	0.3199	0.0329	0.0109	32	0.17	124.02	0
124	Cysteine/Cystine (%)	0682	0.3300	0.0000	0.3199	0.0329	0.0109	32	0.31	124.00	0
124	Cysteine/Cystine (%)	0504	0.3350	0.0100	0.3199	0.0329	0.0109	32	0.46	124.00	0
124	Cysteine/Cystine (%)	0889	0.3393	0.0251	0.3199	0.0329	0.0109	32	0.59	124.99	0
124	Cysteine/Cystine (%)	0684	0.3450	0.0100	0.3199	0.0329	0.0109	32	0.76	124.00	0
124	Cysteine/Cystine (%)	0968	0.3490	0.0000	0.3199	0.0329	0.0109	32	0.89	124.00	0
124	Cysteine/Cystine (%)	0242	0.3500	0.0200	0.3199	0.0329	0.0109	32	0.92	124.05	0
124	Cysteine/Cystine (%)	0610	0.3600	0.0000	0.3199	0.0329	0.0109	32	1.22	124.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
124	Cysteine/Cystine (%)	0619	0.3700	0.0160	0.3199	0.0329	0.0109	32	1.53	124.00	0
124	Cysteine/Cystine (%)	2188	0.3700	0.0000	0.3199	0.0329	0.0109	32	1.53	124.05	0
124	Cysteine/Cystine (%)	0407	0.4070	0.0294	0.3199	0.0329	0.0109	32	2.65	124.05	0
124	Cysteine/Cystine (%)	0675	0.4100	0.0200	0.3199	0.0329	0.0109	32	2.74	124.00	0
124	Cysteine/Cystine (%)	0148	0.4000	0.0600	0.3199	0.0329	0.0109	32	2.44	124.05	1
125	Glutamic (%)	0675	3.370	0.0200	3.683	0.1255	0.0452	33	-2.49	125.00	0
125	Glutamic (%)	0353	3.440	0.0000	3.683	0.1255	0.0452	33	-1.93	125.00	0
125	Glutamic (%)	0652	3.565	0.0500	3.683	0.1255	0.0452	33	-0.94	125.00	0
125	Glutamic (%)	0910	3.570	0.2000	3.683	0.1255	0.0452	33	-0.90	125.00	0
125	Glutamic (%)	0407	3.575	0.0320	3.683	0.1255	0.0452	33	-0.86	125.05	0
125	Glutamic (%)	2059	3.577	0.0030	3.683	0.1255	0.0452	33	-0.85	125.00	0
125	Glutamic (%)	2196	3.609	0.0000	3.683	0.1255	0.0452	33	-0.59	125.05	0
125	Glutamic (%)	0504	3.610	0.0800	3.683	0.1255	0.0452	33	-0.58	125.00	0
125	Glutamic (%)	0872	3.622	0.0600	3.683	0.1255	0.0452	33	-0.48	125.00	0
125	Glutamic (%)	0571	3.625	0.0410	3.683	0.1255	0.0452	33	-0.46	125.00	0
125	Glutamic (%)	0868	3.625	0.0160	3.683	0.1255	0.0452	33	-0.46	125.00	0
125	Glutamic (%)	0644	3.627	0.0230	3.683	0.1255	0.0452	33	-0.45	125.00	0
125	Glutamic (%)	0227	3.630	0.0200	3.683	0.1255	0.0452	33	-0.42	125.00	0
125	Glutamic (%)	0354	3.635	0.0350	3.683	0.1255	0.0452	33	-0.38	125.00	0
125	Glutamic (%)	0098	3.640	0.0420	3.683	0.1255	0.0452	33	-0.34	125.02	0
125	Glutamic (%)	0610	3.640	0.0800	3.683	0.1255	0.0452	33	-0.34	125.05	0
125	Glutamic (%)	0684	3.645	0.0300	3.683	0.1255	0.0452	33	-0.30	125.00	0
125	Glutamic (%)	0941	3.660	0.0000	3.683	0.1255	0.0452	33	-0.18	125.00	0
125	Glutamic (%)	0878	3.664	0.0200	3.683	0.1255	0.0452	33	-0.15	125.00	0
125	Glutamic (%)	0968	3.669	0.0000	3.683	0.1255	0.0452	33	-0.11	125.00	0
125	Glutamic (%)	0859	3.676	0.0120	3.683	0.1255	0.0452	33	-0.05	125.00	0
125	Glutamic (%)	0619	3.690	0.0400	3.683	0.1255	0.0452	33	0.06	125.00	0
125	Glutamic (%)	0918	3.733	0.0135	3.683	0.1255	0.0452	33	0.40	125.00	0
125	Glutamic (%)	0676	3.740	0.0400	3.683	0.1255	0.0452	33	0.46	125.05	0
125	Glutamic (%)	0682	3.752	0.0000	3.683	0.1255	0.0452	33	0.55	125.00	0
125	Glutamic (%)	0870	3.768	0.0543	3.683	0.1255	0.0452	33	0.68	125.00	0
125	Glutamic (%)	0242	3.780	0.0600	3.683	0.1255	0.0452	33	0.78	125.05	0
125	Glutamic (%)	2188	3.860	0.0000	3.683	0.1255	0.0452	33	1.41	125.05	0
125	Glutamic (%)	0889	3.903	0.0989	3.683	0.1255	0.0452	33	1.76	125.99	0
125	Glutamic (%)	0171	3.951	0.2020	3.683	0.1255	0.0452	33	2.14	125.00	0
125	Glutamic (%)	2246	3.965	0.0900	3.683	0.1255	0.0452	33	2.25	125.05	0
125	Glutamic (%)	0148	3.970	0.0200	3.683	0.1255	0.0452	33	2.29	125.05	0
125	Glutamic (%)	0226	4.369	0.1104	3.683	0.1255	0.0452	33	5.47	125.00	0
125	Glutamic (%)	0626	3.817	0.2480	3.683	0.1255	0.0452	33	1.07	125.05	1
126	Glycine (%)	0353	1.260	0.0000	1.359	0.0478	0.0274	33	-2.07	126.00	0
126	Glycine (%)	0407	1.289	0.0180	1.359	0.0478	0.0274	33	-1.46	126.05	0
126	Glycine (%)	0148	1.295	0.0900	1.359	0.0478	0.0274	33	-1.34	126.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
126	Glycine (%)	0675	1.300	0.0200	1.359	0.0478	0.0274	33	-1.23	126.00	0
126	Glycine (%)	0918	1.306	0.0613	1.359	0.0478	0.0274	33	-1.12	126.00	0
126	Glycine (%)	0910	1.310	0.0600	1.359	0.0478	0.0274	33	-1.02	126.00	0
126	Glycine (%)	2188	1.315	0.0300	1.359	0.0478	0.0274	33	-0.92	126.05	0
126	Glycine (%)	0652	1.320	0.0000	1.359	0.0478	0.0274	33	-0.81	126.00	0
126	Glycine (%)	2059	1.340	0.0100	1.359	0.0478	0.0274	33	-0.39	126.00	0
126	Glycine (%)	0619	1.340	0.0000	1.359	0.0478	0.0274	33	-0.39	126.00	0
126	Glycine (%)	0870	1.343	0.0493	1.359	0.0478	0.0274	33	-0.34	126.00	0
126	Glycine (%)	0872	1.343	0.0140	1.359	0.0478	0.0274	33	-0.33	126.00	0
126	Glycine (%)	0644	1.347	0.0210	1.359	0.0478	0.0274	33	-0.26	126.00	0
126	Glycine (%)	0504	1.350	0.0400	1.359	0.0478	0.0274	33	-0.19	126.00	0
126	Glycine (%)	0571	1.354	0.0010	1.359	0.0478	0.0274	33	-0.11	126.00	0
126	Glycine (%)	0227	1.355	0.0100	1.359	0.0478	0.0274	33	-0.08	126.00	0
126	Glycine (%)	0868	1.358	0.0040	1.359	0.0478	0.0274	33	-0.02	126.00	0
126	Glycine (%)	0878	1.358	0.0180	1.359	0.0478	0.0274	33	-0.02	126.00	0
126	Glycine (%)	0098	1.365	0.0080	1.359	0.0478	0.0274	33	0.13	126.02	0
126	Glycine (%)	0610	1.370	0.0000	1.359	0.0478	0.0274	33	0.23	126.05	0
126	Glycine (%)	0354	1.371	0.0150	1.359	0.0478	0.0274	33	0.24	126.00	0
126	Glycine (%)	0968	1.376	0.0000	1.359	0.0478	0.0274	33	0.36	126.00	0
126	Glycine (%)	0859	1.377	0.0049	1.359	0.0478	0.0274	33	0.37	126.00	0
126	Glycine (%)	0242	1.380	0.0800	1.359	0.0478	0.0274	33	0.44	126.05	0
126	Glycine (%)	2196	1.380	0.0000	1.359	0.0478	0.0274	33	0.44	126.05	0
126	Glycine (%)	0941	1.382	0.0500	1.359	0.0478	0.0274	33	0.48	126.00	0
126	Glycine (%)	0171	1.389	0.0760	1.359	0.0478	0.0274	33	0.63	126.00	0
126	Glycine (%)	0676	1.415	0.0100	1.359	0.0478	0.0274	33	1.17	126.05	0
126	Glycine (%)	2246	1.415	0.0100	1.359	0.0478	0.0274	33	1.17	126.05	0
126	Glycine (%)	0682	1.424	0.0000	1.359	0.0478	0.0274	33	1.36	126.00	0
126	Glycine (%)	0684	1.430	0.0400	1.359	0.0478	0.0274	33	1.49	126.00	0
126	Glycine (%)	0626	1.457	0.0830	1.359	0.0478	0.0274	33	2.04	126.05	0
126	Glycine (%)	0226	1.775	0.0815	1.359	0.0478	0.0274	33	8.71	126.00	0
126	Glycine (%)	0889	0.4514	0.0003	1.359	0.0478	0.0274	33	-18.98	126.99	2
127	Histidine (%)	0407	0.4170	0.0200	0.5320	0.0305	0.0118	33	-3.77	127.05	0
127	Histidine (%)	0675	0.4500	0.0200	0.5320	0.0305	0.0118	33	-2.69	127.00	0
127	Histidine (%)	0889	0.4514	0.0003	0.5320	0.0305	0.0118	33	-2.64	127.99	0
127	Histidine (%)	0353	0.4850	0.0100	0.5320	0.0305	0.0118	33	-1.54	127.00	0
127	Histidine (%)	2188	0.4850	0.0100	0.5320	0.0305	0.0118	33	-1.54	127.05	0
127	Histidine (%)	0626	0.5070	0.0400	0.5320	0.0305	0.0118	33	-0.82	127.05	0
127	Histidine (%)	0968	0.5125	0.0030	0.5320	0.0305	0.0118	33	-0.64	127.00	0
127	Histidine (%)	0148	0.5150	0.0300	0.5320	0.0305	0.0118	33	-0.56	127.05	0
127	Histidine (%)	0098	0.5190	0.0020	0.5320	0.0305	0.0118	33	-0.43	127.02	0
127	Histidine (%)	0910	0.5200	0.0200	0.5320	0.0305	0.0118	33	-0.39	127.00	0
127	Histidine (%)	0354	0.5230	0.0080	0.5320	0.0305	0.0118	33	-0.29	127.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
127	Histidine (%)	0571	0.5230	0.0020	0.5320	0.0305	0.0118	33	-0.29	127.00	0
127	Histidine (%)	0868	0.5275	0.0110	0.5320	0.0305	0.0118	33	-0.15	127.00	0
127	Histidine (%)	0859	0.5287	0.0025	0.5320	0.0305	0.0118	33	-0.11	127.00	0
127	Histidine (%)	2059	0.5315	0.0090	0.5320	0.0305	0.0118	33	-0.02	127.00	0
127	Histidine (%)	0872	0.5320	0.0080	0.5320	0.0305	0.0118	33	0.00	127.00	0
127	Histidine (%)	0227	0.5350	0.0100	0.5320	0.0305	0.0118	33	0.10	127.00	0
127	Histidine (%)	0610	0.5350	0.0100	0.5320	0.0305	0.0118	33	0.10	127.05	0
127	Histidine (%)	0870	0.5359	0.0118	0.5320	0.0305	0.0118	33	0.13	127.00	0
127	Histidine (%)	0878	0.5380	0.0060	0.5320	0.0305	0.0118	33	0.20	127.00	0
127	Histidine (%)	0644	0.5390	0.0040	0.5320	0.0305	0.0118	33	0.23	127.00	0
127	Histidine (%)	0171	0.5400	0.0240	0.5320	0.0305	0.0118	33	0.26	127.00	0
127	Histidine (%)	0242	0.5400	0.0200	0.5320	0.0305	0.0118	33	0.26	127.05	0
127	Histidine (%)	0918	0.5436	0.0004	0.5320	0.0305	0.0118	33	0.38	127.00	0
127	Histidine (%)	0504	0.5450	0.0100	0.5320	0.0305	0.0118	33	0.43	127.00	0
127	Histidine (%)	0684	0.5550	0.0100	0.5320	0.0305	0.0118	33	0.75	127.00	0
127	Histidine (%)	0682	0.5560	0.0000	0.5320	0.0305	0.0118	33	0.79	127.00	0
127	Histidine (%)	0676	0.5600	0.0200	0.5320	0.0305	0.0118	33	0.92	127.05	0
127	Histidine (%)	0652	0.5650	0.0300	0.5320	0.0305	0.0118	33	1.08	127.00	0
127	Histidine (%)	2246	0.5650	0.0100	0.5320	0.0305	0.0118	33	1.08	127.05	0
127	Histidine (%)	2196	0.5920	0.0000	0.5320	0.0305	0.0118	33	1.97	127.05	0
127	Histidine (%)	0619	0.6215	0.0010	0.5320	0.0305	0.0118	33	2.93	127.00	0
127	Histidine (%)	0226	0.6594	0.0277	0.5320	0.0305	0.0118	33	4.17	127.00	0
127	Histidine (%)	0941	0.5395	0.1090	0.5320	0.0305	0.0118	33	0.25	127.00	1
128	Isoleucine (%)	0353	0.6550	0.0100	0.8277	0.0513	0.0167	33	-3.37	128.00	0
128	Isoleucine (%)	0675	0.6900	0.0200	0.8277	0.0513	0.0167	33	-2.69	128.00	0
128	Isoleucine (%)	0889	0.7351	0.0042	0.8277	0.0513	0.0167	33	-1.81	128.99	0
128	Isoleucine (%)	0918	0.7545	0.0035	0.8277	0.0513	0.0167	33	-1.43	128.00	0
128	Isoleucine (%)	0868	0.7625	0.0170	0.8277	0.0513	0.0167	33	-1.27	128.00	0
128	Isoleucine (%)	0354	0.7670	0.0080	0.8277	0.0513	0.0167	33	-1.18	128.00	0
128	Isoleucine (%)	0872	0.8025	0.0230	0.8277	0.0513	0.0167	33	-0.49	128.00	0
128	Isoleucine (%)	0682	0.8050	0.0000	0.8277	0.0513	0.0167	33	-0.44	128.00	0
128	Isoleucine (%)	0407	0.8080	0.0720	0.8277	0.0513	0.0167	33	-0.38	128.05	0
128	Isoleucine (%)	0870	0.8106	0.0458	0.8277	0.0513	0.0167	33	-0.33	128.00	0
128	Isoleucine (%)	0652	0.8150	0.0100	0.8277	0.0513	0.0167	33	-0.25	128.00	0
128	Isoleucine (%)	0910	0.8150	0.0500	0.8277	0.0513	0.0167	33	-0.25	128.00	0
128	Isoleucine (%)	0171	0.8160	0.0220	0.8277	0.0513	0.0167	33	-0.23	128.00	0
128	Isoleucine (%)	0619	0.8195	0.0010	0.8277	0.0513	0.0167	33	-0.16	128.00	0
128	Isoleucine (%)	0242	0.8200	0.0200	0.8277	0.0513	0.0167	33	-0.15	128.05	0
128	Isoleucine (%)	0968	0.8235	0.0030	0.8277	0.0513	0.0167	33	-0.08	128.00	0
128	Isoleucine (%)	0878	0.8245	0.0030	0.8277	0.0513	0.0167	33	-0.06	128.00	0
128	Isoleucine (%)	0610	0.8250	0.0100	0.8277	0.0513	0.0167	33	-0.05	128.05	0
128	Isoleucine (%)	0098	0.8300	0.0020	0.8277	0.0513	0.0167	33	0.04	128.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
128	Isoleucine (%)	0571	0.8345	0.0150	0.8277	0.0513	0.0167	33	0.13	128.00	0
128	Isoleucine (%)	0626	0.8370	0.0400	0.8277	0.0513	0.0167	33	0.18	128.05	0
128	Isoleucine (%)	0859	0.8379	0.0056	0.8277	0.0513	0.0167	33	0.20	128.00	0
128	Isoleucine (%)	0227	0.8500	0.0000	0.8277	0.0513	0.0167	33	0.43	128.00	0
128	Isoleucine (%)	2059	0.8535	0.0070	0.8277	0.0513	0.0167	33	0.50	128.00	0
128	Isoleucine (%)	0644	0.8595	0.0050	0.8277	0.0513	0.0167	33	0.62	128.00	0
128	Isoleucine (%)	0676	0.8600	0.0400	0.8277	0.0513	0.0167	33	0.63	128.05	0
128	Isoleucine (%)	0941	0.8710	0.0420	0.8277	0.0513	0.0167	33	0.84	128.00	0
128	Isoleucine (%)	2246	0.8750	0.0100	0.8277	0.0513	0.0167	33	0.92	128.05	0
128	Isoleucine (%)	0504	0.8850	0.0100	0.8277	0.0513	0.0167	33	1.12	128.00	0
128	Isoleucine (%)	2196	0.8910	0.0000	0.8277	0.0513	0.0167	33	1.23	128.05	0
128	Isoleucine (%)	0148	0.9000	0.0200	0.8277	0.0513	0.0167	33	1.41	128.05	0
128	Isoleucine (%)	2188	0.9350	0.0100	0.8277	0.0513	0.0167	33	2.09	128.05	0
128	Isoleucine (%)	0226	0.9378	0.0218	0.8277	0.0513	0.0167	33	2.15	128.00	0
128	Isoleucine (%)	0684	0.7750	0.1100	0.8277	0.0513	0.0167	33	-1.03	128.00	1
129	Leucine (%)	0889	1.341	0.0057	1.636	0.0510	0.0193	33	-5.78	129.99	0
129	Leucine (%)	0675	1.410	0.0200	1.636	0.0510	0.0193	33	-4.43	129.00	0
129	Leucine (%)	0353	1.460	0.0200	1.636	0.0510	0.0193	33	-3.45	129.00	0
129	Leucine (%)	0407	1.479	0.0170	1.636	0.0510	0.0193	33	-3.09	129.05	0
129	Leucine (%)	0910	1.580	0.0600	1.636	0.0510	0.0193	33	-1.10	129.00	0
129	Leucine (%)	0171	1.583	0.0350	1.636	0.0510	0.0193	33	-1.05	129.00	0
129	Leucine (%)	0870	1.597	0.0253	1.636	0.0510	0.0193	33	-0.77	129.00	0
129	Leucine (%)	0652	1.610	0.0200	1.636	0.0510	0.0193	33	-0.51	129.00	0
129	Leucine (%)	0098	1.611	0.0170	1.636	0.0510	0.0193	33	-0.50	129.02	0
129	Leucine (%)	0918	1.616	0.0091	1.636	0.0510	0.0193	33	-0.39	129.00	0
129	Leucine (%)	2059	1.620	0.0030	1.636	0.0510	0.0193	33	-0.32	129.00	0
129	Leucine (%)	0682	1.620	0.0000	1.636	0.0510	0.0193	33	-0.31	129.00	0
129	Leucine (%)	0872	1.625	0.0370	1.636	0.0510	0.0193	33	-0.23	129.00	0
129	Leucine (%)	0878	1.627	0.0230	1.636	0.0510	0.0193	33	-0.19	129.00	0
129	Leucine (%)	0619	1.630	0.0000	1.636	0.0510	0.0193	33	-0.12	129.00	0
129	Leucine (%)	2196	1.633	0.0000	1.636	0.0510	0.0193	33	-0.06	129.05	0
129	Leucine (%)	0644	1.644	0.0030	1.636	0.0510	0.0193	33	0.15	129.00	0
129	Leucine (%)	0242	1.645	0.0300	1.636	0.0510	0.0193	33	0.18	129.05	0
129	Leucine (%)	0571	1.646	0.0310	1.636	0.0510	0.0193	33	0.19	129.00	0
129	Leucine (%)	0868	1.646	0.0120	1.636	0.0510	0.0193	33	0.20	129.00	0
129	Leucine (%)	0504	1.650	0.0200	1.636	0.0510	0.0193	33	0.27	129.00	0
129	Leucine (%)	0684	1.650	0.0200	1.636	0.0510	0.0193	33	0.27	129.00	0
129	Leucine (%)	0968	1.651	0.0040	1.636	0.0510	0.0193	33	0.29	129.00	0
129	Leucine (%)	0227	1.660	0.0200	1.636	0.0510	0.0193	33	0.47	129.00	0
129	Leucine (%)	0610	1.660	0.0200	1.636	0.0510	0.0193	33	0.47	129.05	0
129	Leucine (%)	0354	1.670	0.0040	1.636	0.0510	0.0193	33	0.67	129.00	0
129	Leucine (%)	0941	1.673	0.0070	1.636	0.0510	0.0193	33	0.71	129.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
129	Leucine (%)	0859	1.674	0.0006	1.636	0.0510	0.0193	33	0.73	129.00	0
129	Leucine (%)	0676	1.695	0.0500	1.636	0.0510	0.0193	33	1.16	129.05	0
129	Leucine (%)	0626	1.702	0.0830	1.636	0.0510	0.0193	33	1.28	129.05	0
129	Leucine (%)	2188	1.750	0.0000	1.636	0.0510	0.0193	33	2.23	129.05	0
129	Leucine (%)	0226	1.907	0.0293	1.636	0.0510	0.0193	33	5.31	129.00	0
129	Leucine (%)	2246	2.175	0.0100	1.636	0.0510	0.0193	33	10.56	129.05	0
129	Leucine (%)	0148	1.670	0.1000	1.636	0.0510	0.0193	33	0.67	129.05	1
130	L-Lysine (%)	2053	0.9350	0.0700	1.140	0.0550	0.0219	36	-3.72	130.99	0
130	L-Lysine (%)	0939	1.010	0.0400	1.140	0.0550	0.0219	36	-2.36	130.00	0
130	L-Lysine (%)	0848	1.035	0.0100	1.140	0.0550	0.0219	36	-1.90	130.00	0
130	L-Lysine (%)	0353	1.040	0.0200	1.140	0.0550	0.0219	36	-1.81	130.00	0
130	L-Lysine (%)	0171	1.056	0.0130	1.140	0.0550	0.0219	36	-1.53	130.00	0
130	L-Lysine (%)	0407	1.073	0.0270	1.140	0.0550	0.0219	36	-1.22	130.05	0
130	L-Lysine (%)	0656	1.100	0.0000	1.140	0.0550	0.0219	36	-0.72	130.99	0
130	L-Lysine (%)	0868	1.113	0.0120	1.140	0.0550	0.0219	36	-0.48	130.00	0
130	L-Lysine (%)	0652	1.115	0.0100	1.140	0.0550	0.0219	36	-0.45	130.00	0
130	L-Lysine (%)	0918	1.117	0.0074	1.140	0.0550	0.0219	36	-0.42	130.00	0
130	L-Lysine (%)	0872	1.120	0.0090	1.140	0.0550	0.0219	36	-0.36	130.00	0
130	L-Lysine (%)	0610	1.120	0.0000	1.140	0.0550	0.0219	36	-0.36	130.05	0
130	L-Lysine (%)	0619	1.125	0.0100	1.140	0.0550	0.0219	36	-0.26	130.00	0
130	L-Lysine (%)	0354	1.130	0.0130	1.140	0.0550	0.0219	36	-0.18	130.00	0
130	L-Lysine (%)	0910	1.130	0.0600	1.140	0.0550	0.0219	36	-0.17	130.00	0
130	L-Lysine (%)	0571	1.131	0.0010	1.140	0.0550	0.0219	36	-0.16	130.00	0
130	L-Lysine (%)	0644	1.133	0.0070	1.140	0.0550	0.0219	36	-0.13	130.00	0
130	L-Lysine (%)	0878	1.136	0.0000	1.140	0.0550	0.0219	36	-0.06	130.00	0
130	L-Lysine (%)	2059	1.138	0.0070	1.140	0.0550	0.0219	36	-0.04	130.00	0
130	L-Lysine (%)	2246	1.145	0.0300	1.140	0.0550	0.0219	36	0.10	130.05	0
130	L-Lysine (%)	0098	1.147	0.0270	1.140	0.0550	0.0219	36	0.13	130.02	0
130	L-Lysine (%)	0859	1.153	0.0080	1.140	0.0550	0.0219	36	0.24	130.00	0
130	L-Lysine (%)	0626	1.156	0.0850	1.140	0.0550	0.0219	36	0.29	130.05	0
130	L-Lysine (%)	0870	1.157	0.0207	1.140	0.0550	0.0219	36	0.31	130.00	0
130	L-Lysine (%)	0684	1.160	0.0200	1.140	0.0550	0.0219	36	0.37	130.00	0
130	L-Lysine (%)	0941	1.169	0.0770	1.140	0.0550	0.0219	36	0.53	130.00	0
130	L-Lysine (%)	2188	1.170	0.0000	1.140	0.0550	0.0219	36	0.55	130.05	0
130	L-Lysine (%)	0504	1.180	0.0200	1.140	0.0550	0.0219	36	0.74	130.00	0
130	L-Lysine (%)	0676	1.180	0.0000	1.140	0.0550	0.0219	36	0.74	130.05	0
130	L-Lysine (%)	0968	1.186	0.0100	1.140	0.0550	0.0219	36	0.85	130.00	0
130	L-Lysine (%)	2196	1.193	0.0000	1.140	0.0550	0.0219	36	0.97	130.05	0
130	L-Lysine (%)	0148	1.195	0.0900	1.140	0.0550	0.0219	36	1.01	130.05	0
130	L-Lysine (%)	0242	1.205	0.0100	1.140	0.0550	0.0219	36	1.19	130.05	0
130	L-Lysine (%)	0682	1.233	0.0000	1.140	0.0550	0.0219	36	1.70	130.00	0
130	L-Lysine (%)	0226	1.273	0.0435	1.140	0.0550	0.0219	36	2.42	130.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
130	L-Lysine (%)	0227	1.275	0.0300	1.140	0.0550	0.0219	36	2.46	130.00	0
130	L-Lysine (%)	0889	1.105	0.1500	1.140	0.0550	0.0219	36	-0.63	130.99	1
130	L-Lysine (%)	0014	1.120	0.1600	1.140	0.0550	0.0219	36	-0.36	130.05	1
130	L-Lysine (%)	0675	2.050	0.0200	1.140	0.0550	0.0219	36	16.56	130.00	2
131	Methionine (%)	0918	0.3000	0.0000	0.4108	0.0278	0.0133	37	-3.99	131.00	0
131	Methionine (%)	2246	0.3300	0.0200	0.4108	0.0278	0.0133	37	-2.91	131.05	0
131	Methionine (%)	0148	0.3450	0.0500	0.4108	0.0278	0.0133	37	-2.37	131.05	0
131	Methionine (%)	0682	0.3670	0.0000	0.4108	0.0278	0.0133	37	-1.58	131.00	0
131	Methionine (%)	0626	0.3715	0.0110	0.4108	0.0278	0.0133	37	-1.42	131.05	0
131	Methionine (%)	0014	0.3785	0.0510	0.4108	0.0278	0.0133	37	-1.17	131.05	0
131	Methionine (%)	0675	0.3800	0.0200	0.4108	0.0278	0.0133	37	-1.11	131.00	0
131	Methionine (%)	0939	0.3950	0.0300	0.4108	0.0278	0.0133	37	-0.57	131.00	0
131	Methionine (%)	0354	0.3990	0.0020	0.4108	0.0278	0.0133	37	-0.43	131.00	0
131	Methionine (%)	0652	0.4000	0.0200	0.4108	0.0278	0.0133	37	-0.39	131.00	0
131	Methionine (%)	0968	0.4005	0.0230	0.4108	0.0278	0.0133	37	-0.37	131.00	0
131	Methionine (%)	0872	0.4025	0.0150	0.4108	0.0278	0.0133	37	-0.30	131.00	0
131	Methionine (%)	0171	0.4045	0.0030	0.4108	0.0278	0.0133	37	-0.23	131.00	0
131	Methionine (%)	0848	0.4050	0.0100	0.4108	0.0278	0.0133	37	-0.21	131.00	0
131	Methionine (%)	0941	0.4050	0.0000	0.4108	0.0278	0.0133	37	-0.21	131.00	0
131	Methionine (%)	0878	0.4075	0.0070	0.4108	0.0278	0.0133	37	-0.12	131.00	0
131	Methionine (%)	0227	0.4100	0.0000	0.4108	0.0278	0.0133	37	-0.03	131.00	0
131	Methionine (%)	0870	0.4110	0.0098	0.4108	0.0278	0.0133	37	0.01	131.00	0
131	Methionine (%)	0098	0.4115	0.0070	0.4108	0.0278	0.0133	37	0.02	131.02	0
131	Methionine (%)	2059	0.4135	0.0050	0.4108	0.0278	0.0133	37	0.10	131.00	0
131	Methionine (%)	0859	0.4143	0.0015	0.4108	0.0278	0.0133	37	0.12	131.00	0
131	Methionine (%)	0619	0.4145	0.0070	0.4108	0.0278	0.0133	37	0.13	131.00	0
131	Methionine (%)	0504	0.4150	0.0100	0.4108	0.0278	0.0133	37	0.15	131.00	0
131	Methionine (%)	0610	0.4150	0.0100	0.4108	0.0278	0.0133	37	0.15	131.05	0
131	Methionine (%)	0910	0.4200	0.0200	0.4108	0.0278	0.0133	37	0.33	131.00	0
131	Methionine (%)	0868	0.4225	0.0030	0.4108	0.0278	0.0133	37	0.42	131.00	0
131	Methionine (%)	0571	0.4230	0.0120	0.4108	0.0278	0.0133	37	0.44	131.00	0
131	Methionine (%)	0353	0.4250	0.0100	0.4108	0.0278	0.0133	37	0.51	131.00	0
131	Methionine (%)	0242	0.4300	0.0000	0.4108	0.0278	0.0133	37	0.69	131.05	0
131	Methionine (%)	0676	0.4300	0.0200	0.4108	0.0278	0.0133	37	0.69	131.05	0
131	Methionine (%)	0684	0.4350	0.0100	0.4108	0.0278	0.0133	37	0.87	131.00	0
131	Methionine (%)	0644	0.4370	0.0020	0.4108	0.0278	0.0133	37	0.94	131.00	0
131	Methionine (%)	0889	0.4400	0.0200	0.4108	0.0278	0.0133	37	1.05	131.99	0
131	Methionine (%)	0407	0.4508	0.0223	0.4108	0.0278	0.0133	37	1.44	131.05	0
131	Methionine (%)	2196	0.4540	0.0000	0.4108	0.0278	0.0133	37	1.56	131.05	0
131	Methionine (%)	0226	0.4940	0.0502	0.4108	0.0278	0.0133	37	3.00	131.00	0
131	Methionine (%)	2188	0.4950	0.0100	0.4108	0.0278	0.0133	37	3.03	131.05	0
132	Phenylalanine (%)	0407	0.8260	0.0060	0.9653	0.0338	0.0185	33	-4.12	132.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
132	Phenylalanine (%)	0675	0.8800	0.0200	0.9653	0.0338	0.0185	33	-2.52	132.00	0
132	Phenylalanine (%)	2188	0.9050	0.0100	0.9653	0.0338	0.0185	33	-1.78	132.05	0
132	Phenylalanine (%)	0889	0.9232	0.0045	0.9653	0.0338	0.0185	33	-1.25	132.99	0
132	Phenylalanine (%)	0918	0.9239	0.0222	0.9653	0.0338	0.0185	33	-1.23	132.00	0
132	Phenylalanine (%)	0652	0.9300	0.0000	0.9653	0.0338	0.0185	33	-1.04	132.00	0
132	Phenylalanine (%)	0910	0.9400	0.0400	0.9653	0.0338	0.0185	33	-0.75	132.00	0
132	Phenylalanine (%)	0171	0.9430	0.0180	0.9653	0.0338	0.0185	33	-0.66	132.00	0
132	Phenylalanine (%)	0098	0.9490	0.0080	0.9653	0.0338	0.0185	33	-0.48	132.02	0
132	Phenylalanine (%)	2059	0.9505	0.0010	0.9653	0.0338	0.0185	33	-0.44	132.00	0
132	Phenylalanine (%)	0684	0.9550	0.0300	0.9653	0.0338	0.0185	33	-0.31	132.00	0
132	Phenylalanine (%)	0870	0.9585	0.0164	0.9653	0.0338	0.0185	33	-0.20	132.00	0
132	Phenylalanine (%)	0682	0.9590	0.0000	0.9653	0.0338	0.0185	33	-0.19	132.00	0
132	Phenylalanine (%)	0227	0.9600	0.0200	0.9653	0.0338	0.0185	33	-0.16	132.00	0
132	Phenylalanine (%)	0878	0.9605	0.0010	0.9653	0.0338	0.0185	33	-0.14	132.00	0
132	Phenylalanine (%)	0354	0.9645	0.0010	0.9653	0.0338	0.0185	33	-0.03	132.00	0
132	Phenylalanine (%)	0242	0.9650	0.0300	0.9653	0.0338	0.0185	33	-0.01	132.05	0
132	Phenylalanine (%)	0644	0.9660	0.0060	0.9653	0.0338	0.0185	33	0.02	132.00	0
132	Phenylalanine (%)	0941	0.9660	0.0100	0.9653	0.0338	0.0185	33	0.02	132.00	0
132	Phenylalanine (%)	0872	0.9670	0.0240	0.9653	0.0338	0.0185	33	0.05	132.00	0
132	Phenylalanine (%)	0571	0.9710	0.0120	0.9653	0.0338	0.0185	33	0.17	132.00	0
132	Phenylalanine (%)	0353	0.9750	0.0100	0.9653	0.0338	0.0185	33	0.29	132.00	0
132	Phenylalanine (%)	0968	0.9775	0.0030	0.9653	0.0338	0.0185	33	0.36	132.00	0
132	Phenylalanine (%)	0626	0.9845	0.0670	0.9653	0.0338	0.0185	33	0.57	132.05	0
132	Phenylalanine (%)	2246	0.9850	0.0100	0.9653	0.0338	0.0185	33	0.58	132.05	0
132	Phenylalanine (%)	0619	0.9915	0.0370	0.9653	0.0338	0.0185	33	0.77	132.00	0
132	Phenylalanine (%)	0504	0.9950	0.0100	0.9653	0.0338	0.0185	33	0.88	132.00	0
132	Phenylalanine (%)	0859	0.9999	0.0066	0.9653	0.0338	0.0185	33	1.02	132.00	0
132	Phenylalanine (%)	0610	1.000	0.0400	0.9653	0.0338	0.0185	33	1.02	132.05	0
132	Phenylalanine (%)	2196	1.004	0.0000	0.9653	0.0338	0.0185	33	1.14	132.05	0
132	Phenylalanine (%)	0676	1.035	0.0500	0.9653	0.0338	0.0185	33	2.06	132.05	0
132	Phenylalanine (%)	0868	1.045	0.0120	0.9653	0.0338	0.0185	33	2.35	132.00	0
132	Phenylalanine (%)	0226	1.106	0.0839	0.9653	0.0338	0.0185	33	4.14	132.00	0
132	Phenylalanine (%)	0148	0.9300	0.1400	0.9653	0.0338	0.0185	33	-1.04	132.05	1
133	Proline (%)	0407	1.258	0.0170	1.374	0.0607	0.0325	33	-1.92	133.05	0
133	Proline (%)	0870	1.288	0.0117	1.374	0.0607	0.0325	33	-1.42	133.00	0
133	Proline (%)	0910	1.290	0.0600	1.374	0.0607	0.0325	33	-1.39	133.00	0
133	Proline (%)	2196	1.298	0.0000	1.374	0.0607	0.0325	33	-1.26	133.05	0
133	Proline (%)	0619	1.300	0.0200	1.374	0.0607	0.0325	33	-1.22	133.00	0
133	Proline (%)	0504	1.325	0.0100	1.374	0.0607	0.0325	33	-0.81	133.00	0
133	Proline (%)	0878	1.329	0.0180	1.374	0.0607	0.0325	33	-0.74	133.00	0
133	Proline (%)	0571	1.339	0.0010	1.374	0.0607	0.0325	33	-0.59	133.00	0
133	Proline (%)	0242	1.340	0.0800	1.374	0.0607	0.0325	33	-0.56	133.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
133	Proline (%)	0682	1.350	0.0000	1.374	0.0607	0.0325	33	-0.40	133.00	0
133	Proline (%)	0872	1.360	0.0380	1.374	0.0607	0.0325	33	-0.23	133.00	0
133	Proline (%)	0889	1.362	0.0552	1.374	0.0607	0.0325	33	-0.20	133.99	0
133	Proline (%)	0227	1.365	0.0300	1.374	0.0607	0.0325	33	-0.15	133.00	0
133	Proline (%)	0353	1.365	0.0500	1.374	0.0607	0.0325	33	-0.15	133.00	0
133	Proline (%)	0675	1.370	0.0200	1.374	0.0607	0.0325	33	-0.07	133.00	0
133	Proline (%)	0148	1.370	0.0200	1.374	0.0607	0.0325	33	-0.07	133.05	0
133	Proline (%)	0859	1.371	0.0308	1.374	0.0607	0.0325	33	-0.05	133.00	0
133	Proline (%)	2059	1.371	0.0060	1.374	0.0607	0.0325	33	-0.05	133.00	0
133	Proline (%)	0610	1.375	0.0100	1.374	0.0607	0.0325	33	0.01	133.05	0
133	Proline (%)	0171	1.377	0.0620	1.374	0.0607	0.0325	33	0.05	133.00	0
133	Proline (%)	0354	1.380	0.0040	1.374	0.0607	0.0325	33	0.10	133.00	0
133	Proline (%)	0941	1.382	0.0300	1.374	0.0607	0.0325	33	0.13	133.00	0
133	Proline (%)	0676	1.385	0.0700	1.374	0.0607	0.0325	33	0.18	133.05	0
133	Proline (%)	0652	1.390	0.1400	1.374	0.0607	0.0325	33	0.26	133.00	0
133	Proline (%)	0644	1.396	0.0230	1.374	0.0607	0.0325	33	0.35	133.00	0
133	Proline (%)	0868	1.397	0.0020	1.374	0.0607	0.0325	33	0.38	133.00	0
133	Proline (%)	0684	1.425	0.0100	1.374	0.0607	0.0325	33	0.84	133.00	0
133	Proline (%)	0626	1.453	0.0640	1.374	0.0607	0.0325	33	1.30	133.05	0
133	Proline (%)	0918	1.459	0.0215	1.374	0.0607	0.0325	33	1.39	133.00	0
133	Proline (%)	0968	1.460	0.0150	1.374	0.0607	0.0325	33	1.40	133.00	0
133	Proline (%)	2246	1.515	0.0100	1.374	0.0607	0.0325	33	2.32	133.05	0
133	Proline (%)	2188	1.520	0.0000	1.374	0.0607	0.0325	33	2.40	133.05	0
133	Proline (%)	0226	1.607	0.1431	1.374	0.0607	0.0325	33	3.84	133.00	0
134	Serine (%)	0889	0.8708	0.0020	1.007	0.0491	0.0147	32	-2.77	134.99	0
134	Serine (%)	0504	0.8750	0.0500	1.007	0.0491	0.0147	32	-2.69	134.00	0
134	Serine (%)	0918	0.9463	0.0052	1.007	0.0491	0.0147	32	-1.23	134.00	0
134	Serine (%)	0644	0.9505	0.0130	1.007	0.0491	0.0147	32	-1.15	134.00	0
134	Serine (%)	2059	0.9540	0.0040	1.007	0.0491	0.0147	32	-1.08	134.00	0
134	Serine (%)	0148	0.9550	0.0100	1.007	0.0491	0.0147	32	-1.06	134.05	0
134	Serine (%)	0407	0.9625	0.0010	1.007	0.0491	0.0147	32	-0.90	134.05	0
134	Serine (%)	0910	0.9750	0.0300	1.007	0.0491	0.0147	32	-0.65	134.00	0
134	Serine (%)	0610	0.9750	0.0100	1.007	0.0491	0.0147	32	-0.65	134.05	0
134	Serine (%)	0227	0.9800	0.0200	1.007	0.0491	0.0147	32	-0.55	134.00	0
134	Serine (%)	0353	0.9900	0.0000	1.007	0.0491	0.0147	32	-0.34	134.00	0
134	Serine (%)	0675	0.9900	0.0200	1.007	0.0491	0.0147	32	-0.34	134.00	0
134	Serine (%)	0098	0.9900	0.0120	1.007	0.0491	0.0147	32	-0.34	134.02	0
134	Serine (%)	0652	1.000	0.0000	1.007	0.0491	0.0147	32	-0.14	134.00	0
134	Serine (%)	0968	1.005	0.0030	1.007	0.0491	0.0147	32	-0.05	134.00	0
134	Serine (%)	0571	1.011	0.0080	1.007	0.0491	0.0147	32	0.08	134.00	0
134	Serine (%)	0878	1.011	0.0080	1.007	0.0491	0.0147	32	0.08	134.00	0
134	Serine (%)	0872	1.015	0.0150	1.007	0.0491	0.0147	32	0.15	134.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
134	Serine (%)	0619	1.015	0.0100	1.007	0.0491	0.0147	32	0.16	134.00	0
134	Serine (%)	0941	1.018	0.0290	1.007	0.0491	0.0147	32	0.22	134.00	0
134	Serine (%)	2196	1.023	0.0000	1.007	0.0491	0.0147	32	0.33	134.05	0
134	Serine (%)	0859	1.023	0.0010	1.007	0.0491	0.0147	32	0.34	134.00	0
134	Serine (%)	2188	1.030	0.0200	1.007	0.0491	0.0147	32	0.47	134.05	0
134	Serine (%)	0868	1.037	0.0070	1.007	0.0491	0.0147	32	0.60	134.00	0
134	Serine (%)	0354	1.037	0.0060	1.007	0.0491	0.0147	32	0.61	134.00	0
134	Serine (%)	0242	1.040	0.0200	1.007	0.0491	0.0147	32	0.67	134.05	0
134	Serine (%)	0626	1.048	0.0730	1.007	0.0491	0.0147	32	0.83	134.05	0
134	Serine (%)	0870	1.069	0.0063	1.007	0.0491	0.0147	32	1.26	134.00	0
134	Serine (%)	0682	1.071	0.0000	1.007	0.0491	0.0147	32	1.31	134.00	0
134	Serine (%)	0676	1.075	0.0700	1.007	0.0491	0.0147	32	1.39	134.05	0
134	Serine (%)	0226	1.106	0.0090	1.007	0.0491	0.0147	32	2.02	134.00	0
134	Serine (%)	0171	1.116	0.0070	1.007	0.0491	0.0147	32	2.21	134.00	0
134	Serine (%)	0684	1.060	0.1400	1.007	0.0491	0.0147	32	1.08	134.00	1
135	Threonine (%)	0407	0.6200	0.0280	0.7917	0.0351	0.0151	33	-4.89	135.05	0
135	Threonine (%)	2188	0.6200	0.0000	0.7917	0.0351	0.0151	33	-4.89	135.05	0
135	Threonine (%)	0889	0.6364	0.0023	0.7917	0.0351	0.0151	33	-4.42	135.99	0
135	Threonine (%)	0353	0.7450	0.0100	0.7917	0.0351	0.0151	33	-1.33	135.00	0
135	Threonine (%)	0910	0.7500	0.0400	0.7917	0.0351	0.0151	33	-1.19	135.00	0
135	Threonine (%)	0626	0.7565	0.0370	0.7917	0.0351	0.0151	33	-1.00	135.05	0
135	Threonine (%)	0652	0.7600	0.0200	0.7917	0.0351	0.0151	33	-0.90	135.00	0
135	Threonine (%)	0148	0.7650	0.0100	0.7917	0.0351	0.0151	33	-0.76	135.05	0
135	Threonine (%)	2059	0.7725	0.0010	0.7917	0.0351	0.0151	33	-0.55	135.00	0
135	Threonine (%)	0504	0.7800	0.0200	0.7917	0.0351	0.0151	33	-0.33	135.00	0
135	Threonine (%)	0675	0.7800	0.0200	0.7917	0.0351	0.0151	33	-0.33	135.00	0
135	Threonine (%)	0644	0.7820	0.0100	0.7917	0.0351	0.0151	33	-0.28	135.00	0
135	Threonine (%)	0868	0.7845	0.0030	0.7917	0.0351	0.0151	33	-0.21	135.00	0
135	Threonine (%)	0571	0.7900	0.0060	0.7917	0.0351	0.0151	33	-0.05	135.00	0
135	Threonine (%)	0610	0.7900	0.0200	0.7917	0.0351	0.0151	33	-0.05	135.05	0
135	Threonine (%)	0878	0.7920	0.0080	0.7917	0.0351	0.0151	33	0.01	135.00	0
135	Threonine (%)	0098	0.7920	0.0240	0.7917	0.0351	0.0151	33	0.01	135.02	0
135	Threonine (%)	0872	0.7935	0.0170	0.7917	0.0351	0.0151	33	0.05	135.00	0
135	Threonine (%)	0227	0.7950	0.0300	0.7917	0.0351	0.0151	33	0.09	135.00	0
135	Threonine (%)	0859	0.7987	0.0016	0.7917	0.0351	0.0151	33	0.20	135.00	0
135	Threonine (%)	0682	0.8030	0.0000	0.7917	0.0351	0.0151	33	0.32	135.00	0
135	Threonine (%)	0619	0.8045	0.0110	0.7917	0.0351	0.0151	33	0.36	135.00	0
135	Threonine (%)	0941	0.8060	0.0300	0.7917	0.0351	0.0151	33	0.41	135.00	0
135	Threonine (%)	0870	0.8084	0.0117	0.7917	0.0351	0.0151	33	0.47	135.00	0
135	Threonine (%)	0354	0.8140	0.0120	0.7917	0.0351	0.0151	33	0.63	135.00	0
135	Threonine (%)	0918	0.8145	0.0014	0.7917	0.0351	0.0151	33	0.65	135.00	0
135	Threonine (%)	0968	0.8200	0.0100	0.7917	0.0351	0.0151	33	0.80	135.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
135	Threonine (%)	0676	0.8200	0.0400	0.7917	0.0351	0.0151	33	0.80	135.05	0
135	Threonine (%)	2196	0.8250	0.0000	0.7917	0.0351	0.0151	33	0.95	135.05	0
135	Threonine (%)	0242	0.8350	0.0100	0.7917	0.0351	0.0151	33	1.23	135.05	0
135	Threonine (%)	0684	0.8500	0.0200	0.7917	0.0351	0.0151	33	1.66	135.00	0
135	Threonine (%)	0226	0.8643	0.0350	0.7917	0.0351	0.0151	33	2.06	135.00	0
135	Threonine (%)	2246	0.8850	0.0100	0.7917	0.0351	0.0151	33	2.65	135.05	0
135	Threonine (%)	0171	0.9920	0.0940	0.7917	0.0351	0.0151	33	5.70	135.00	1
136	Tryptophan (%)	2246	0.1050	0.0100	0.2440	0.0361	0.0096	23	-3.86	136.05	0
136	Tryptophan (%)	0407	0.1295	0.0410	0.2440	0.0361	0.0096	23	-3.18	136.05	0
136	Tryptophan (%)	0918	0.1900	0.0200	0.2440	0.0361	0.0096	23	-1.50	136.00	0
136	Tryptophan (%)	0171	0.1905	0.0070	0.2440	0.0361	0.0096	23	-1.48	136.00	0
136	Tryptophan (%)	0504	0.2250	0.0100	0.2440	0.0361	0.0096	23	-0.53	136.99	0
136	Tryptophan (%)	0859	0.2305	0.0091	0.2440	0.0361	0.0096	23	-0.38	136.03	0
136	Tryptophan (%)	0684	0.2350	0.0100	0.2440	0.0361	0.0096	23	-0.25	136.00	0
136	Tryptophan (%)	0878	0.2360	0.0060	0.2440	0.0361	0.0096	23	-0.22	136.01	0
136	Tryptophan (%)	0868	0.2385	0.0030	0.2440	0.0361	0.0096	23	-0.15	136.01	0
136	Tryptophan (%)	0910	0.2400	0.0000	0.2440	0.0361	0.0096	23	-0.11	136.00	0
136	Tryptophan (%)	0098	0.2425	0.0050	0.2440	0.0361	0.0096	23	-0.04	136.02	0
136	Tryptophan (%)	0619	0.2470	0.0080	0.2440	0.0361	0.0096	23	0.08	136.00	0
136	Tryptophan (%)	0242	0.2500	0.0400	0.2440	0.0361	0.0096	23	0.17	136.05	0
136	Tryptophan (%)	0872	0.2505	0.0070	0.2440	0.0361	0.0096	23	0.18	136.03	0
136	Tryptophan (%)	2059	0.2545	0.0010	0.2440	0.0361	0.0096	23	0.29	136.03	0
136	Tryptophan (%)	0644	0.2580	0.0000	0.2440	0.0361	0.0096	23	0.39	136.01	0
136	Tryptophan (%)	0571	0.2625	0.0010	0.2440	0.0361	0.0096	23	0.51	136.03	0
136	Tryptophan (%)	0870	0.2650	0.0132	0.2440	0.0361	0.0096	23	0.58	136.00	0
136	Tryptophan (%)	0148	0.2650	0.0100	0.2440	0.0361	0.0096	23	0.58	136.03	0
136	Tryptophan (%)	0682	0.2710	0.0000	0.2440	0.0361	0.0096	23	0.75	136.00	0
136	Tryptophan (%)	0227	0.2850	0.0100	0.2440	0.0361	0.0096	23	1.14	136.00	0
136	Tryptophan (%)	0353	0.3050	0.0100	0.2440	0.0361	0.0096	23	1.69	136.03	0
136	Tryptophan (%)	0889	0.4523	0.0002	0.2440	0.0361	0.0096	23	5.78	136.99	0
137	Tyrosine (%)	0889	0.4523	0.0002	0.6436	0.0864	0.0147	27	-2.21	137.99	0
137	Tyrosine (%)	0910	0.5150	0.0100	0.6436	0.0864	0.0147	27	-1.49	137.00	0
137	Tyrosine (%)	0682	0.5180	0.0000	0.6436	0.0864	0.0147	27	-1.45	137.00	0
137	Tyrosine (%)	0407	0.5395	0.0050	0.6436	0.0864	0.0147	27	-1.20	137.05	0
137	Tyrosine (%)	0626	0.5675	0.0470	0.6436	0.0864	0.0147	27	-0.88	137.05	0
137	Tyrosine (%)	0675	0.5700	0.0200	0.6436	0.0864	0.0147	27	-0.85	137.00	0
137	Tyrosine (%)	0227	0.5900	0.0400	0.6436	0.0864	0.0147	27	-0.62	137.00	0
137	Tyrosine (%)	0870	0.5926	0.0204	0.6436	0.0864	0.0147	27	-0.59	137.00	0
137	Tyrosine (%)	0171	0.6015	0.0090	0.6436	0.0864	0.0147	27	-0.49	137.00	0
137	Tyrosine (%)	0941	0.6060	0.0500	0.6436	0.0864	0.0147	27	-0.43	137.00	0
137	Tyrosine (%)	0918	0.6337	0.0270	0.6436	0.0864	0.0147	27	-0.11	137.00	0
137	Tyrosine (%)	2246	0.6450	0.0100	0.6436	0.0864	0.0147	27	0.02	137.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
137	Tyrosine (%)	0872	0.6475	0.0050	0.6436	0.0864	0.0147	27	0.05	137.00	0
137	Tyrosine (%)	0354	0.6560	0.0040	0.6436	0.0864	0.0147	27	0.14	137.00	0
137	Tyrosine (%)	0968	0.6595	0.0010	0.6436	0.0864	0.0147	27	0.18	137.00	0
137	Tyrosine (%)	0684	0.6650	0.0300	0.6436	0.0864	0.0147	27	0.25	137.00	0
137	Tyrosine (%)	0242	0.6650	0.0300	0.6436	0.0864	0.0147	27	0.25	137.05	0
137	Tyrosine (%)	0644	0.6740	0.0040	0.6436	0.0864	0.0147	27	0.35	137.00	0
137	Tyrosine (%)	0504	0.6750	0.0100	0.6436	0.0864	0.0147	27	0.36	137.00	0
137	Tyrosine (%)	0098	0.6790	0.0040	0.6436	0.0864	0.0147	27	0.41	137.02	0
137	Tyrosine (%)	2188	0.7000	0.0000	0.6436	0.0864	0.0147	27	0.65	137.05	0
137	Tyrosine (%)	0353	0.7100	0.0000	0.6436	0.0864	0.0147	27	0.77	137.00	0
137	Tyrosine (%)	2196	0.7270	0.0000	0.6436	0.0864	0.0147	27	0.96	137.05	0
137	Tyrosine (%)	2059	0.7395	0.0070	0.6436	0.0864	0.0147	27	1.11	137.00	0
137	Tyrosine (%)	0610	0.7400	0.0200	0.6436	0.0864	0.0147	27	1.12	137.05	0
137	Tyrosine (%)	0226	0.7899	0.0246	0.6436	0.0864	0.0147	27	1.69	137.00	0
137	Tyrosine (%)	0676	0.8000	0.0200	0.6436	0.0864	0.0147	27	1.81	137.05	0
137	Tyrosine (%)	0148	0.7300	0.1000	0.6436	0.0864	0.0147	27	1.00	137.05	1
138	Valine (%)	0353	0.7850	0.0100	0.9898	0.0528	0.0213	32	-3.88	138.00	0
138	Valine (%)	0889	0.8794	0.0010	0.9898	0.0528	0.0213	32	-2.09	138.99	0
138	Valine (%)	0868	0.8970	0.0220	0.9898	0.0528	0.0213	32	-1.76	138.00	0
138	Valine (%)	0870	0.9358	0.0499	0.9898	0.0528	0.0213	32	-1.02	138.00	0
138	Valine (%)	0226	0.9378	0.0218	0.9898	0.0528	0.0213	32	-0.99	138.00	0
138	Valine (%)	0407	0.9420	0.0920	0.9898	0.0528	0.0213	32	-0.91	138.05	0
138	Valine (%)	0171	0.9430	0.0140	0.9898	0.0528	0.0213	32	-0.89	138.00	0
138	Valine (%)	0652	0.9450	0.0100	0.9898	0.0528	0.0213	32	-0.85	138.00	0
138	Valine (%)	0354	0.9520	0.0100	0.9898	0.0528	0.0213	32	-0.72	138.00	0
138	Valine (%)	2196	0.9580	0.0000	0.9898	0.0528	0.0213	32	-0.60	138.05	0
138	Valine (%)	0872	0.9655	0.0370	0.9898	0.0528	0.0213	32	-0.46	138.00	0
138	Valine (%)	0918	0.9743	0.0018	0.9898	0.0528	0.0213	32	-0.29	138.00	0
138	Valine (%)	0910	0.9750	0.1100	0.9898	0.0528	0.0213	32	-0.28	138.00	0
138	Valine (%)	0878	0.9895	0.0150	0.9898	0.0528	0.0213	32	-0.01	138.00	0
138	Valine (%)	0571	0.9905	0.0110	0.9898	0.0528	0.0213	32	0.01	138.00	0
138	Valine (%)	0968	0.9905	0.0010	0.9898	0.0528	0.0213	32	0.01	138.00	0
138	Valine (%)	0682	0.9910	0.0000	0.9898	0.0528	0.0213	32	0.02	138.00	0
138	Valine (%)	0619	0.9935	0.0130	0.9898	0.0528	0.0213	32	0.07	138.00	0
138	Valine (%)	0242	0.9950	0.0300	0.9898	0.0528	0.0213	32	0.10	138.05	0
138	Valine (%)	0626	1.005	0.0620	0.9898	0.0528	0.0213	32	0.29	138.05	0
138	Valine (%)	0859	1.007	0.0074	0.9898	0.0528	0.0213	32	0.33	138.00	0
138	Valine (%)	0098	1.013	0.0090	0.9898	0.0528	0.0213	32	0.43	138.02	0
138	Valine (%)	2059	1.018	0.0050	0.9898	0.0528	0.0213	32	0.53	138.00	0
138	Valine (%)	0941	1.027	0.0560	0.9898	0.0528	0.0213	32	0.71	138.00	0
138	Valine (%)	0644	1.035	0.0030	0.9898	0.0528	0.0213	32	0.85	138.00	0
138	Valine (%)	0227	1.035	0.0100	0.9898	0.0528	0.0213	32	0.86	138.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
138	Valine (%)	0148	1.040	0.0200	0.9898	0.0528	0.0213	32	0.95	138.05	0
138	Valine (%)	2188	1.040	0.0000	0.9898	0.0528	0.0213	32	0.95	138.05	0
138	Valine (%)	0504	1.055	0.0100	0.9898	0.0528	0.0213	32	1.24	138.00	0
138	Valine (%)	0610	1.060	0.0200	0.9898	0.0528	0.0213	32	1.33	138.05	0
138	Valine (%)	0676	1.060	0.0200	0.9898	0.0528	0.0213	32	1.33	138.05	0
138	Valine (%)	2246	1.115	0.0100	0.9898	0.0528	0.0213	32	2.37	138.05	0
138	Valine (%)	0684	0.9600	0.1200	0.9898	0.0528	0.0213	32	-0.56	138.00	1
138	Valine (%)	0675	0.3400	0.0200	0.9898	0.0528	0.0213	32	-12.32	138.00	2
139	Taurine (%)	0148	0.0100	0.0000	0.0792	0.0949	0.0007	4		139.05	0
139	Taurine (%)	0889	0.0156	0.0006	0.0792	0.0949	0.0007	4		139.99	0
139	Taurine (%)	0682	0.0770	0.0000	0.0792	0.0949	0.0007	4		139.00	0
139	Taurine (%)	0407	0.2140	0.0020	0.0792	0.0949	0.0007	4		139.05	0
139	Taurine (%)	0504	0.1400	0.0200	0.0792	0.0949	0.0007	4		139.00	1
139	Taurine (%)	0098	< 0.01		0.0792	0.0949	0.0007	4		139.02	5
160	Fructose (%)	0297	0.1085	0.0350				3		160.99	0
160	Fructose (%)	0148	0.2350	0.0100				3		160.99	0
160	Fructose (%)	0227	0.3050	0.0100				3		160.99	0
161	Galactose (%)	0297	0.0000	0.0000				0		161.99	4
162	Glucose (%)	0297	0.1525	0.0170				3		162.99	0
162	Glucose (%)	0227	0.2200	0.0600				3		162.99	0
162	Glucose (%)	0148	0.2550	0.1100				3		162.99	0
163	Lactose (%)	0148	0.0000	0.0000				0		163.99	4
163	Lactose (%)	0297	0.0000	0.0000				0		163.99	4
163	Lactose (%)	0227	< 0.15					0		163.99	5
164	Maltose (%)	0297	0.2025	0.0310				1		164.99	0
164	Maltose (%)	0148	0.0000	0.0000				1		164.99	4
165	Sucrose (%)	0297	1.899	0.1010				3		165.99	0
165	Sucrose (%)	0148	1.975	0.0700				3		165.99	0
165	Sucrose (%)	0227	2.150	0.0800				3		165.99	0
166	Raffinose (%)	0227	0.5000	0.0000				2		166.99	0
166	Raffinose (%)	0297	0.5300	0.0040				2		166.99	0
167	Stachyose (%)	0227	0.9350	0.0300				2		167.99	0
167	Stachyose (%)	0297	1.062	0.0740				2		167.99	0
354	Decoquinatate (mg/kg (ppm))	0006	29.00	0.0000				1		354.02	0
365	Monensin (mg/kg (ppm))	2113	0.4050	0.1300				2		365.05	0
365	Monensin (mg/kg (ppm))	0610	0.4700	0.0200				2		365.02	0
379	Salinomycin (mg/kg (ppm))	2113	120.0	0.0000				1		379.05	0
391	Narasin (mg/kg (ppm))	2113	0.3450	0.0100	95.82	19.11	2.447	13	-5.00	391.03	0
391	Narasin (mg/kg (ppm))	2135	78.27	2.097	95.82	19.11	2.447	13	-0.92	391.00	0
391	Narasin (mg/kg (ppm))	2055	82.00	6.000	95.82	19.11	2.447	13	-0.72	391.03	0
391	Narasin (mg/kg (ppm))	0033	84.00	0.6000	95.82	19.11	2.447	13	-0.62	391.00	0
391	Narasin (mg/kg (ppm))	0910	90.35	0.7000	95.82	19.11	2.447	13	-0.29	391.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
391	Narasin (mg/kg (ppm))	0668	92.05	0.1000	95.82	19.11	2.447	13	-0.20	391.00	0
391	Narasin (mg/kg (ppm))	0668	93.25	0.9000	95.82	19.11	2.447	13	-0.13	391.99	0
391	Narasin (mg/kg (ppm))	0218	95.72	1.342	95.82	19.11	2.447	13	-0.01	391.00	0
391	Narasin (mg/kg (ppm))	0038	101.7	5.930	95.82	19.11	2.447	13	0.31	391.03	0
391	Narasin (mg/kg (ppm))	0610	104.2	1.030	95.82	19.11	2.447	13	0.44	391.03	0
391	Narasin (mg/kg (ppm))	0610	108.0	2.000	95.82	19.11	2.447	13	0.64	391.01	0
391	Narasin (mg/kg (ppm))	2234	137.5	3.200	95.82	19.11	2.447	13	2.18	391.03	0
391	Narasin (mg/kg (ppm))	0941	141.0	7.900	95.82	19.11	2.447	13	2.36	391.03	0
400	Water Activity (Units)	0843	0.5490	0.0000	0.5648	0.0159	0.0020	10	-1.00	400.99	0
400	Water Activity (Units)	0407	0.5518	0.0014	0.5648	0.0159	0.0020	10	-0.82	400.01	0
400	Water Activity (Units)	2033	0.5540	0.0020	0.5648	0.0159	0.0020	10	-0.68	400.99	0
400	Water Activity (Units)	0942	0.5580	0.0080	0.5648	0.0159	0.0020	10	-0.43	400.01	0
400	Water Activity (Units)	2181	0.5581	0.0006	0.5648	0.0159	0.0020	10	-0.42	400.01	0
400	Water Activity (Units)	0843	0.5637	0.0013	0.5648	0.0159	0.0020	10	-0.07	400.01	0
400	Water Activity (Units)	0589	0.5670	0.0020	0.5648	0.0159	0.0020	10	0.14	400.01	0
400	Water Activity (Units)	0668	0.5700	0.0000	0.5648	0.0159	0.0020	10	0.33	400.01	0
400	Water Activity (Units)	2073	0.5880	0.0040	0.5648	0.0159	0.0020	10	1.46	400.99	0
400	Water Activity (Units)	2109	0.5921	0.0009	0.5648	0.0159	0.0020	10	1.71	400.01	0
412	Starch, Dietary (%)	0353	26.47	0.1300				1		412.01	0
516	Arsenic, Total (mg / kg (ppm))	0910	0.1000	0.0200	0.1327	0.0382	0.0212	8	-0.86	516.52	0
516	Arsenic, Total (mg / kg (ppm))	0045	0.1090	0.0120	0.1327	0.0382	0.0212	8	-0.62	516.00	0
516	Arsenic, Total (mg / kg (ppm))	0904	0.1100	0.0600	0.1327	0.0382	0.0212	8	-0.60	516.00	0
516	Arsenic, Total (mg / kg (ppm))	0171	0.1155	0.0050	0.1327	0.0382	0.0212	8	-0.45	516.00	0
516	Arsenic, Total (mg / kg (ppm))	0553	0.1235	0.0050	0.1327	0.0382	0.0212	8	-0.24	516.53	0
516	Arsenic, Total (mg / kg (ppm))	0227	0.1350	0.0280	0.1327	0.0382	0.0212	8	0.06	516.53	0
516	Arsenic, Total (mg / kg (ppm))	0918	0.1787	0.0136	0.1327	0.0382	0.0212	8	1.21	516.53	0
516	Arsenic, Total (mg / kg (ppm))	0407	1.114	0.0257	0.1327	0.0382	0.0212	8	25.72	516.43	0
518	Cadmium (mg / kg (ppm))	0695	0.0845	0.0050	0.1088	0.0073	0.0066	9	-3.33	518.41	0
518	Cadmium (mg / kg (ppm))	0968	0.1020	0.0020	0.1088	0.0073	0.0066	9	-0.94	518.43	0
518	Cadmium (mg / kg (ppm))	0910	0.1050	0.0100	0.1088	0.0073	0.0066	9	-0.52	518.52	0
518	Cadmium (mg / kg (ppm))	0407	0.1088	0.0105	0.1088	0.0073	0.0066	9	-0.01	518.43	0
518	Cadmium (mg / kg (ppm))	0407	0.1089	0.0015	0.1088	0.0073	0.0066	9	0.00	518.41	0
518	Cadmium (mg / kg (ppm))	0918	0.1117	0.0052	0.1088	0.0073	0.0066	9	0.39	518.53	0
518	Cadmium (mg / kg (ppm))	0553	0.1125	0.0030	0.1088	0.0073	0.0066	9	0.50	518.53	0
518	Cadmium (mg / kg (ppm))	0171	0.1130	0.0020	0.1088	0.0073	0.0066	9	0.57	518.41	0
518	Cadmium (mg / kg (ppm))	2246	0.2300	0.0200	0.1088	0.0073	0.0066	9	16.61	518.34	0
518	Cadmium (mg / kg (ppm))	0227	0.1315	0.0390	0.1088	0.0073	0.0066	9	3.11	518.53	1
518	Cadmium (mg / kg (ppm))	0689	< 0.2		0.1088	0.0073	0.0066	9		518.31	5
520	Chromium (mg / kg (ppm))	0968	2.181	0.1000	4.778	0.9975	0.2694	12	-2.60	520.43	0
520	Chromium (mg / kg (ppm))	0918	2.437	0.1407	4.778	0.9975	0.2694	12	-2.35	520.53	0
520	Chromium (mg / kg (ppm))	0407	3.995	0.0100	4.778	0.9975	0.2694	12	-0.79	520.41	0
520	Chromium (mg / kg (ppm))	0685	4.370	0.0200	4.778	0.9975	0.2694	12	-0.41	520.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
520	Chromium (mg / kg (ppm))	0171	4.561	0.1560	4.778	0.9975	0.2694	12	-0.22	520.41	0
520	Chromium (mg / kg (ppm))	0045	5.070	0.2200	4.778	0.9975	0.2694	12	0.29	520.42	0
520	Chromium (mg / kg (ppm))	2053	5.235	0.0900	4.778	0.9975	0.2694	12	0.46	520.42	0
520	Chromium (mg / kg (ppm))	0164	5.260	0.3600	4.778	0.9975	0.2694	12	0.48	520.53	0
520	Chromium (mg / kg (ppm))	0407	5.298	0.2050	4.778	0.9975	0.2694	12	0.52	520.43	0
520	Chromium (mg / kg (ppm))	0693	5.451	1.121	4.778	0.9975	0.2694	12	0.67	520.42	0
520	Chromium (mg / kg (ppm))	0510	5.745	0.1100	4.778	0.9975	0.2694	12	0.97	520.43	0
520	Chromium (mg / kg (ppm))	0553	5.790	0.7000	4.778	0.9975	0.2694	12	1.01	520.53	0
526	Lead (mg / kg (ppm))	0171	0.1275	0.0150	0.2040	0.0547	0.0100	8	-1.40	526.41	0
526	Lead (mg / kg (ppm))	0227	0.1825	0.0090	0.2040	0.0547	0.0100	8	-0.39	526.53	0
526	Lead (mg / kg (ppm))	0910	0.1850	0.0100	0.2040	0.0547	0.0100	8	-0.35	526.52	0
526	Lead (mg / kg (ppm))	0553	0.1880	0.0100	0.2040	0.0547	0.0100	8	-0.29	526.53	0
526	Lead (mg / kg (ppm))	0918	0.1909	0.0034	0.2040	0.0547	0.0100	8	-0.24	526.53	0
526	Lead (mg / kg (ppm))	0407	0.2223	0.0126	0.2040	0.0547	0.0100	8	0.33	526.41	0
526	Lead (mg / kg (ppm))	2152	0.2500	0.0200	0.2040	0.0547	0.0100	8	0.84	526.41	0
526	Lead (mg / kg (ppm))	2246	0.9700	0.0000	0.2040	0.0547	0.0100	8	14.00	526.34	0
526	Lead (mg / kg (ppm))	0407	0.2436	0.1683	0.2040	0.0547	0.0100	8	0.72	526.43	1
526	Lead (mg / kg (ppm))	0689	< 0.2		0.2040	0.0547	0.0100	8		526.31	5
526	Lead (mg / kg (ppm))	0904	< 0.3		0.2040	0.0547	0.0100	8		526.34	5
539	Nickel (mg / kg (ppm))	0407	3.040	0.0435	3.597	0.5712	0.1910	6	-0.98	539.41	0
539	Nickel (mg / kg (ppm))	0407	3.231	0.3856	3.597	0.5712	0.1910	6	-0.64	539.43	0
539	Nickel (mg / kg (ppm))	0941	3.386	0.0000	3.597	0.5712	0.1910	6	-0.37	539.43	0
539	Nickel (mg / kg (ppm))	0171	3.677	0.0110	3.597	0.5712	0.1910	6	0.14	539.41	0
539	Nickel (mg / kg (ppm))	0553	3.795	0.1900	3.597	0.5712	0.1910	6	0.35	539.53	0
539	Nickel (mg / kg (ppm))	0918	4.699	0.5158	3.597	0.5712	0.1910	6	1.93	539.53	0
702	Butyric Acid (4:0) (%)	0619	0.0035	0.0010				1		702.00	0
704	Caproic Acid (6:0) (%)	0619	0.0015	0.0010				1		704.00	0
706	Caprylic acid (8:0) (%) (w/w)	0619	0.0000	0.0000				0		706.01	4
708	Capric acid (10:0) (%) (w/w)	0619	0.0010	0.0000				1		708.01	0
710	Lauric Acid (12:0) (%) (w/w)	0676	0.0000	0.0000				0		710.99	4
710	Lauric Acid (12:0) (%) (w/w)	0619	0.0115	0.0230				0		710.01	4
710	Lauric Acid (12:0) (%) (w/w)	0226	< 0.005					0		710.99	5
710	Lauric Acid (12:0) (%) (w/w)	0164	< 0.02					0		710.99	5
714	Myristic Acid (14:0) (%) (w/w)	0226	0.0210	0.0000				3		714.99	0
714	Myristic Acid (14:0) (%) (w/w)	0164	0.0300	0.0000				3		714.99	0
714	Myristic Acid (14:0) (%) (w/w)	0619	0.0340	0.0000				3		714.01	0
716	Palmitic Acid (16:0) (%) (w/w)	0226	0.7510	0.0080				3		716.99	0
716	Palmitic Acid (16:0) (%) (w/w)	0164	0.7700	0.0000				3		716.99	0
716	Palmitic Acid (16:0) (%) (w/w)	0619	0.8930	0.0060				3		716.01	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0676	0.0340	0.0000	0.0398	0.0049	0.0000	4		718.99	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0226	0.0390	0.0000	0.0398	0.0049	0.0000	4		718.99	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0164	0.0400	0.0000	0.0398	0.0049	0.0000	4		718.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
718	Palmitoleic Acid (9c-16:1) (% (w/w))	0619	0.0460	0.0000	0.0398	0.0049	0.0000	4	718.01	0	
722	Stearic Acid (18:0) (% (w/w))	0619	0.2380	0.0020				3	722.01	0	
722	Stearic Acid (18:0) (% (w/w))	0164	0.2600	0.0000				3	722.99	0	
722	Stearic Acid (18:0) (% (w/w))	0226	0.2670	0.0000				3	722.99	0	
724	Oleic Acid (9c-18:1) (% (w/w))	0226	1.047	0.0120				3	724.99	0	
724	Oleic Acid (9c-18:1) (% (w/w))	0164	1.100	0.0000				3	724.99	0	
724	Oleic Acid (9c-18:1) (% (w/w))	0619	1.275	0.0100				3	724.01	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0164	1.405	0.0100	1.540	0.1732	0.0158	5	726.99	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0226	1.418	0.0040	1.540	0.1732	0.0158	5	726.99	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0148	1.425	0.0300	1.540	0.1732	0.0158	5	726.02	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0676	1.676	0.0250	1.540	0.1732	0.0158	5	726.99	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0619	1.775	0.0100	1.540	0.1732	0.0158	5	726.01	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v))	0619	0.0815	0.0010	0.0905	0.0126	0.0010	4	728.01	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v))	0226	0.0820	0.0000	0.0905	0.0126	0.0010	4	728.99	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v))	0164	0.0900	0.0000	0.0905	0.0126	0.0010	4	728.99	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v))	0676	0.1085	0.0030	0.0905	0.0126	0.0010	4	728.99	0	
730	Arachidic Acid (20:0) (% (w/w))	0226	0.0130	0.0000				2	730.99	0	
730	Arachidic Acid (20:0) (% (w/w))	0619	0.0380	0.0000				2	730.01	0	
730	Arachidic Acid (20:0) (% (w/w))	0164	< 0.02					2	730.99	5	
732	Gondoic Acid (11c-20:1) (% (w/w))	0164	0.0200	0.0000				3	732.99	0	
732	Gondoic Acid (11c-20:1) (% (w/w))	0226	0.0240	0.0020				3	732.99	0	
732	Gondoic Acid (11c-20:1) (% (w/w))	0619	0.0405	0.0010				3	732.01	0	
736	Arachidonic Acid (5c,8c,11c,14c-20:4) (% (w/w))	0676	0.0070	0.0020				1	736.99	0	
736	Arachidonic Acid (5c,8c,11c,14c-20:4) (% (w/w))	0619	0.0000	0.0000				1	736.01	4	
738	Mead Acid (11c,14c,17c-20:3) (% (w/w))	0619	0.0000	0.0000				0	738.01	4	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c) (% (w/w))	0619	0.0000	0.0000				0	740.01	4	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c) (% (w/w))	0676	0.0000	0.0000				0	740.99	4	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c) (% (w/w))	0226	< 0.005					0	740.99	5	
742	Behenic Acid (22:0) (% (w/w))	0226	0.0080	0.0000				1	742.99	0	
742	Behenic Acid (22:0) (% (w/w))	0164	< 0.02					1	742.99	5	
744	Erucic Acid (13c-22:1) (% (w/w))	0619	0.0000	0.0000				0	744.01	4	
744	Erucic Acid (13c-22:1) (% (w/w))	0226	< 0.005					0	744.99	5	
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c) (% (w/w))	0619	0.0000	0.0000				0	746.01	4	
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c) (% (w/w))	0676	0.0000	0.0000				0	746.99	4	
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c) (% (w/w))	0226	< 0.005					0	746.99	5	
748	Lignoceric Acid (24:0) (% (w/w))	0226	0.0090	0.0000				1	748.99	0	
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22c) (% (w/w))	0619	0.0000	0.0000				0	750.01	4	
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22c) (% (w/w))	0676	0.0000	0.0000				0	750.99	4	
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22c) (% (w/w))	0226	< 0.005					0	750.99	5	
752	Nervonic Acid (24:1) isomers (% (w/w))	0619	0.0000	0.0000				0	752.01	4	
752	Nervonic Acid (24:1) isomers (% (w/w))	0226	< 0.005					0	752.99	5	
754	Total n-3 Polyunsaturated (Omega-3) Fatty Acids (% (w/w))	0148	0.0900	0.0000				3	754.02	0	

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
754	Total n-3 Polyunsaturated (Omega-3) Fatty Aci	0164	0.0900	0.0000				3	754.99	0	
754	Total n-3 Polyunsaturated (Omega-3) Fatty Aci	0676	0.1100	0.0000				3	754.99	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Aci	0164	1.410	0.0000				3	756.99	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Aci	0148	1.430	0.0400				3	756.01	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Aci	0676	1.695	0.0300				3	756.99	0	
758	Total Saturated Fatty Acids (% (w/w))	0164	1.110	0.0000				1	758.99	0	
762	Total Monounsaturated Fatty Acids (% (w/w))	0164	1.160	0.0000				1	762.99	0	
766	Total Polyunsaturated Fatty Acids (% (w/w))	0164	1.505	0.0100				1	766.99	0	
770	Total Fat (equivalent to NLEA) (% (w/w))	0164	3.965	0.0100				1	770.99	0	
772	Total Fatty Acids (% (w/w))	0226	3.679	0.0260				2	772.99	0	
772	Total Fatty Acids (% (w/w))	0164	3.795	0.0100				2	772.99	0	

Notes: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2 and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = Rejected for duplicates too far apart, 2 = Rejected as outlier, 5 = A reporting limit and 4 = zeros submitted as values. Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). Flag 3 indicates not used in statistics.