



Animal Feed Scheme
Chicken Layer, Medicated
Test Material Code # 202021

Labs Reporting: 170
Analytes Reported 114
Issue Date : 02/29/2020

Analyte Proficiency Testing Report

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
000	Urea (%)	0278	0.3000	0.0000				2		000.02	0
000	Urea (%)	2129	0.4773	0.0974				2		000.99	0
001	Loss on Drying (%)	0618	7.265	0.2900	9.315	0.2964	0.1125	68	-6.92	001.99	0
001	Loss on Drying (%)	0074	8.020	0.0400	9.315	0.2964	0.1125	68	-4.37	001.07	0
001	Loss on Drying (%)	2155	8.160	0.2000	9.315	0.2964	0.1125	68	-3.90	001.99	0
001	Loss on Drying (%)	2081	8.290	0.2200	9.315	0.2964	0.1125	68	-3.46	001.99	0
001	Loss on Drying (%)	0019	8.450	0.1000	9.315	0.2964	0.1125	68	-2.92	001.07	0
001	Loss on Drying (%)	0536	8.505	0.0700	9.315	0.2964	0.1125	68	-2.73	001.99	0
001	Loss on Drying (%)	0889	8.680	0.1600	9.315	0.2964	0.1125	68	-2.14	001.99	0
001	Loss on Drying (%)	0675	8.865	0.2500	9.315	0.2964	0.1125	68	-1.52	001.07	0
001	Loss on Drying (%)	0034	8.870	0.2300	9.315	0.2964	0.1125	68	-1.50	001.07	0
001	Loss on Drying (%)	0676	8.935	0.1100	9.315	0.2964	0.1125	68	-1.28	001.99	0
001	Loss on Drying (%)	0510	8.950	0.1000	9.315	0.2964	0.1125	68	-1.23	001.99	0
001	Loss on Drying (%)	0297	8.970	0.1200	9.315	0.2964	0.1125	68	-1.16	001.07	0
001	Loss on Drying (%)	0852	9.060	0.0200	9.315	0.2964	0.1125	68	-0.86	001.99	0
001	Loss on Drying (%)	0853	9.100	0.0000	9.315	0.2964	0.1125	68	-0.73	001.07	0
001	Loss on Drying (%)	2146	9.105	0.0100	9.315	0.2964	0.1125	68	-0.71	001.07	0
001	Loss on Drying (%)	0148	9.115	0.2300	9.315	0.2964	0.1125	68	-0.67	001.07	0
001	Loss on Drying (%)	0226	9.150	0.1200	9.315	0.2964	0.1125	68	-0.56	001.07	0
001	Loss on Drying (%)	0027	9.155	0.0570	9.315	0.2964	0.1125	68	-0.54	001.00	0
001	Loss on Drying (%)	0644	9.185	0.0700	9.315	0.2964	0.1125	68	-0.44	001.05	0
001	Loss on Drying (%)	0027	9.200	0.0400	9.315	0.2964	0.1125	68	-0.39	001.07	0
001	Loss on Drying (%)	0353	9.225	0.0700	9.315	0.2964	0.1125	68	-0.30	001.07	0
001	Loss on Drying (%)	0171	9.240	0.0600	9.315	0.2964	0.1125	68	-0.25	001.07	0
001	Loss on Drying (%)	0968	9.252	0.0010	9.315	0.2964	0.1125	68	-0.21	001.05	0
001	Loss on Drying (%)	0083	9.255	0.0700	9.315	0.2964	0.1125	68	-0.20	001.07	0
001	Loss on Drying (%)	0505	9.260	0.0400	9.315	0.2964	0.1125	68	-0.19	001.07	0
001	Loss on Drying (%)	2192	9.270	0.0000	9.315	0.2964	0.1125	68	-0.15	001.07	0
001	Loss on Drying (%)	2009	9.275	0.0133	9.315	0.2964	0.1125	68	-0.13	001.07	0
001	Loss on Drying (%)	0066	9.280	0.0000	9.315	0.2964	0.1125	68	-0.12	001.07	0
001	Loss on Drying (%)	0948	9.280	0.0200	9.315	0.2964	0.1125	68	-0.12	001.99	0
001	Loss on Drying (%)	0186	9.287	0.2730	9.315	0.2964	0.1125	68	-0.10	001.99	0
001	Loss on Drying (%)	0345	9.290	0.0800	9.315	0.2964	0.1125	68	-0.08	001.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
001	Loss on Drying (%)	0918	9.290	0.0200	9.315	0.2964	0.1125	68	-0.08	001.99	0
001	Loss on Drying (%)	2053	9.300	0.0000	9.315	0.2964	0.1125	68	-0.05	001.07	0
001	Loss on Drying (%)	0169	9.310	0.0200	9.315	0.2964	0.1125	68	-0.02	001.00	0
001	Loss on Drying (%)	0619	9.335	0.0700	9.315	0.2964	0.1125	68	0.07	001.03	0
001	Loss on Drying (%)	0590	9.340	0.4200	9.315	0.2964	0.1125	68	0.08	001.07	0
001	Loss on Drying (%)	0610	9.350	0.1000	9.315	0.2964	0.1125	68	0.12	001.99	0
001	Loss on Drying (%)	0037	9.355	0.0100	9.315	0.2964	0.1125	68	0.14	001.99	0
001	Loss on Drying (%)	2059	9.375	0.0100	9.315	0.2964	0.1125	68	0.20	001.07	0
001	Loss on Drying (%)	0512	9.383	0.4720	9.315	0.2964	0.1125	68	0.23	001.07	0
001	Loss on Drying (%)	0142	9.385	0.0100	9.315	0.2964	0.1125	68	0.24	001.07	0
001	Loss on Drying (%)	0049	9.390	0.3600	9.315	0.2964	0.1125	68	0.25	001.07	0
001	Loss on Drying (%)	0910	9.400	0.0000	9.315	0.2964	0.1125	68	0.29	001.99	0
001	Loss on Drying (%)	0139	9.410	0.0000	9.315	0.2964	0.1125	68	0.32	001.07	0
001	Loss on Drying (%)	0035	9.415	0.1100	9.315	0.2964	0.1125	68	0.34	001.07	0
001	Loss on Drying (%)	0581	9.445	0.0300	9.315	0.2964	0.1125	68	0.44	001.07	0
001	Loss on Drying (%)	0560	9.450	0.5000	9.315	0.2964	0.1125	68	0.46	001.99	0
001	Loss on Drying (%)	0571	9.475	0.0700	9.315	0.2964	0.1125	68	0.54	001.07	0
001	Loss on Drying (%)	0019	9.475	0.1100	9.315	0.2964	0.1125	68	0.54	001.99	0
001	Loss on Drying (%)	0003	9.500	0.1800	9.315	0.2964	0.1125	68	0.62	001.07	0
001	Loss on Drying (%)	2109	9.510	0.2000	9.315	0.2964	0.1125	68	0.66	001.07	0
001	Loss on Drying (%)	0638	9.530	0.0200	9.315	0.2964	0.1125	68	0.73	001.07	0
001	Loss on Drying (%)	0689	9.550	0.1000	9.315	0.2964	0.1125	68	0.79	001.07	0
001	Loss on Drying (%)	0098	9.555	0.1100	9.315	0.2964	0.1125	68	0.81	001.07	0
001	Loss on Drying (%)	0878	9.560	0.0200	9.315	0.2964	0.1125	68	0.83	001.03	0
001	Loss on Drying (%)	2062	9.570	0.0009	9.315	0.2964	0.1125	68	0.86	001.03	0
001	Loss on Drying (%)	0407	9.584	0.0105	9.315	0.2964	0.1125	68	0.91	001.07	0
001	Loss on Drying (%)	0028	9.585	0.0500	9.315	0.2964	0.1125	68	0.91	001.07	0
001	Loss on Drying (%)	0876	9.600	0.2000	9.315	0.2964	0.1125	68	0.96	001.07	0
001	Loss on Drying (%)	0309	9.605	0.0300	9.315	0.2964	0.1125	68	0.98	001.00	0
001	Loss on Drying (%)	0872	9.615	0.0300	9.315	0.2964	0.1125	68	1.01	001.07	0
001	Loss on Drying (%)	2259	9.686	0.0130	9.315	0.2964	0.1125	68	1.25	001.07	0
001	Loss on Drying (%)	0693	9.723	0.0270	9.315	0.2964	0.1125	68	1.38	001.07	0
001	Loss on Drying (%)	0843	9.725	0.4100	9.315	0.2964	0.1125	68	1.38	001.03	0
001	Loss on Drying (%)	0278	9.725	0.3900	9.315	0.2964	0.1125	68	1.38	001.07	0
001	Loss on Drying (%)	0843	9.725	0.4100	9.315	0.2964	0.1125	68	1.38	001.07	0
001	Loss on Drying (%)	0357	9.753	0.0736	9.315	0.2964	0.1125	68	1.48	001.99	0
001	Loss on Drying (%)	0682	10.38	0.0000	9.315	0.2964	0.1125	68	3.59	001.07	0
001	Loss on Drying (%)	0045	10.13	0.9000	9.315	0.2964	0.1125	68	2.75	001.07	1
001	Loss on Drying (%)	0366	10.20	1.200	9.315	0.2964	0.1125	68	2.99	001.07	1
002	Protein, Crude (%)	0622	17.22	0.0410	18.23	0.2833	0.1484	163	-3.57	002.05	0
002	Protein, Crude (%)	0643	17.32	0.1000	18.23	0.2833	0.1484	163	-3.23	002.99	0
002	Protein, Crude (%)	0536	17.41	0.0100	18.23	0.2833	0.1484	163	-2.93	002.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	2303	17.51	0.5800	18.23	0.2833	0.1484	163	-2.56	002.06	0
002	Protein, Crude (%)	0921	17.60	0.0000	18.23	0.2833	0.1484	163	-2.24	002.05	0
002	Protein, Crude (%)	2199	17.60	0.4000	18.23	0.2833	0.1484	163	-2.24	002.05	0
002	Protein, Crude (%)	2259	17.72	0.0180	18.23	0.2833	0.1484	163	-1.82	002.01	0
002	Protein, Crude (%)	0043	17.73	0.2000	18.23	0.2833	0.1484	163	-1.78	002.01	0
002	Protein, Crude (%)	0003	17.75	0.0700	18.23	0.2833	0.1484	163	-1.73	002.06	0
002	Protein, Crude (%)	0042	17.75	0.1000	18.23	0.2833	0.1484	163	-1.71	002.06	0
002	Protein, Crude (%)	0539	17.78	0.3900	18.23	0.2833	0.1484	163	-1.62	002.06	0
002	Protein, Crude (%)	0638	17.78	0.2600	18.23	0.2833	0.1484	163	-1.60	002.04	0
002	Protein, Crude (%)	0689	17.80	0.0000	18.23	0.2833	0.1484	163	-1.53	002.05	0
002	Protein, Crude (%)	0910	17.85	0.5000	18.23	0.2833	0.1484	163	-1.36	002.01	0
002	Protein, Crude (%)	0939	17.86	0.1600	18.23	0.2833	0.1484	163	-1.32	002.05	0
002	Protein, Crude (%)	2109	17.87	0.2100	18.23	0.2833	0.1484	163	-1.30	002.05	0
002	Protein, Crude (%)	0598	17.89	0.0600	18.23	0.2833	0.1484	163	-1.21	002.06	0
002	Protein, Crude (%)	0139	17.90	0.2700	18.23	0.2833	0.1484	163	-1.20	002.06	0
002	Protein, Crude (%)	0043	17.91	0.4560	18.23	0.2833	0.1484	163	-1.15	002.06	0
002	Protein, Crude (%)	2168	17.92	0.1600	18.23	0.2833	0.1484	163	-1.11	002.06	0
002	Protein, Crude (%)	0653	17.93	0.1800	18.23	0.2833	0.1484	163	-1.07	002.01	0
002	Protein, Crude (%)	0693	17.93	0.2230	18.23	0.2833	0.1484	163	-1.06	002.06	0
002	Protein, Crude (%)	0676	17.94	0.2500	18.23	0.2833	0.1484	163	-1.06	002.06	0
002	Protein, Crude (%)	0948	17.95	0.0300	18.23	0.2833	0.1484	163	-1.02	002.05	0
002	Protein, Crude (%)	0619	17.95	0.1000	18.23	0.2833	0.1484	163	-1.00	002.05	0
002	Protein, Crude (%)	0036	17.95	0.0572	18.23	0.2833	0.1484	163	-0.99	002.02	0
002	Protein, Crude (%)	0870	17.96	0.2769	18.23	0.2833	0.1484	163	-0.98	002.06	0
002	Protein, Crude (%)	0914	17.99	0.0100	18.23	0.2833	0.1484	163	-0.88	002.01	0
002	Protein, Crude (%)	2009	17.99	0.0500	18.23	0.2833	0.1484	163	-0.86	002.05	0
002	Protein, Crude (%)	2129	18.00	0.0980	18.23	0.2833	0.1484	163	-0.84	002.06	0
002	Protein, Crude (%)	0675	18.00	0.1600	18.23	0.2833	0.1484	163	-0.83	002.05	0
002	Protein, Crude (%)	0014	18.00	0.4000	18.23	0.2833	0.1484	163	-0.83	002.06	0
002	Protein, Crude (%)	2053	18.00	0.2000	18.23	0.2833	0.1484	163	-0.83	002.06	0
002	Protein, Crude (%)	0263	18.00	0.0826	18.23	0.2833	0.1484	163	-0.81	002.06	0
002	Protein, Crude (%)	2006	18.01	0.3900	18.23	0.2833	0.1484	163	-0.81	002.05	0
002	Protein, Crude (%)	0028	18.02	0.0000	18.23	0.2833	0.1484	163	-0.75	002.06	0
002	Protein, Crude (%)	0038	18.03	0.0100	18.23	0.2833	0.1484	163	-0.74	002.06	0
002	Protein, Crude (%)	2188	18.04	0.0100	18.23	0.2833	0.1484	163	-0.70	002.01	0
002	Protein, Crude (%)	0941	18.04	0.1500	18.23	0.2833	0.1484	163	-0.70	002.05	0
002	Protein, Crude (%)	0074	18.04	0.2100	18.23	0.2833	0.1484	163	-0.70	002.06	0
002	Protein, Crude (%)	0202	18.04	0.0500	18.23	0.2833	0.1484	163	-0.70	002.06	0
002	Protein, Crude (%)	0169	18.06	0.0200	18.23	0.2833	0.1484	163	-0.61	002.00	0
002	Protein, Crude (%)	0098	18.07	0.0100	18.23	0.2833	0.1484	163	-0.60	002.08	0
002	Protein, Crude (%)	0066	18.07	0.0700	18.23	0.2833	0.1484	163	-0.60	002.02	0
002	Protein, Crude (%)	2304	18.07	0.0330	18.23	0.2833	0.1484	163	-0.59	002.09	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0006	18.08	0.0300	18.23	0.2833	0.1484	163	-0.56	002.06	0
002	Protein, Crude (%)	0870	18.08	0.0370	18.23	0.2833	0.1484	163	-0.55	002.01	0
002	Protein, Crude (%)	2193	18.08	0.1200	18.23	0.2833	0.1484	163	-0.54	002.01	0
002	Protein, Crude (%)	0939	18.09	0.1100	18.23	0.2833	0.1484	163	-0.53	002.06	0
002	Protein, Crude (%)	0682	18.09	0.0000	18.23	0.2833	0.1484	163	-0.51	002.06	0
002	Protein, Crude (%)	0563	18.10	0.2982	18.23	0.2833	0.1484	163	-0.47	002.08	0
002	Protein, Crude (%)	0017	18.10	0.0000	18.23	0.2833	0.1484	163	-0.47	002.06	0
002	Protein, Crude (%)	0098	18.10	0.2000	18.23	0.2833	0.1484	163	-0.47	002.06	0
002	Protein, Crude (%)	0278	18.10	0.0000	18.23	0.2833	0.1484	163	-0.47	002.06	0
002	Protein, Crude (%)	0529	18.10	0.4000	18.23	0.2833	0.1484	163	-0.47	002.06	0
002	Protein, Crude (%)	0541	18.11	0.0100	18.23	0.2833	0.1484	163	-0.45	002.06	0
002	Protein, Crude (%)	0646	18.12	0.1200	18.23	0.2833	0.1484	163	-0.40	002.06	0
002	Protein, Crude (%)	0354	18.12	0.0400	18.23	0.2833	0.1484	163	-0.40	002.05	0
002	Protein, Crude (%)	2150	18.14	0.2400	18.23	0.2833	0.1484	163	-0.33	002.06	0
002	Protein, Crude (%)	0297	18.15	0.0700	18.23	0.2833	0.1484	163	-0.31	002.06	0
002	Protein, Crude (%)	0652	18.15	0.1000	18.23	0.2833	0.1484	163	-0.30	002.01	0
002	Protein, Crude (%)	0010	18.15	0.1000	18.23	0.2833	0.1484	163	-0.30	002.06	0
002	Protein, Crude (%)	0049	18.15	0.1000	18.23	0.2833	0.1484	163	-0.30	002.06	0
002	Protein, Crude (%)	0626	18.16	0.0420	18.23	0.2833	0.1484	163	-0.27	002.06	0
002	Protein, Crude (%)	0407	18.17	0.2710	18.23	0.2833	0.1484	163	-0.24	002.06	0
002	Protein, Crude (%)	2196	18.17	0.0000	18.23	0.2833	0.1484	163	-0.23	002.01	0
002	Protein, Crude (%)	2062	18.18	0.0076	18.23	0.2833	0.1484	163	-0.19	002.05	0
002	Protein, Crude (%)	2192	18.19	0.0700	18.23	0.2833	0.1484	163	-0.17	002.05	0
002	Protein, Crude (%)	0100	18.19	0.2100	18.23	0.2833	0.1484	163	-0.17	002.06	0
002	Protein, Crude (%)	0505	18.19	0.1100	18.23	0.2833	0.1484	163	-0.17	002.06	0
002	Protein, Crude (%)	0035	18.19	0.0300	18.23	0.2833	0.1484	163	-0.17	002.06	0
002	Protein, Crude (%)	0897	18.20	0.0000	18.23	0.2833	0.1484	163	-0.12	002.05	0
002	Protein, Crude (%)	0853	18.20	0.4000	18.23	0.2833	0.1484	163	-0.12	002.06	0
002	Protein, Crude (%)	0619	18.20	0.2000	18.23	0.2833	0.1484	163	-0.12	002.06	0
002	Protein, Crude (%)	0956	18.20	0.2000	18.23	0.2833	0.1484	163	-0.12	002.06	0
002	Protein, Crude (%)	0878	18.21	0.1200	18.23	0.2833	0.1484	163	-0.08	002.05	0
002	Protein, Crude (%)	0011	18.21	0.3600	18.23	0.2833	0.1484	163	-0.08	002.06	0
002	Protein, Crude (%)	0881	18.21	0.0290	18.23	0.2833	0.1484	163	-0.08	002.05	0
002	Protein, Crude (%)	0358	18.22	0.0200	18.23	0.2833	0.1484	163	-0.05	002.06	0
002	Protein, Crude (%)	0964	18.22	0.1000	18.23	0.2833	0.1484	163	-0.04	002.06	0
002	Protein, Crude (%)	0968	18.23	0.0400	18.23	0.2833	0.1484	163	-0.02	002.01	0
002	Protein, Crude (%)	0889	18.24	0.1100	18.23	0.2833	0.1484	163	0.00	002.11	0
002	Protein, Crude (%)	0354	18.24	0.0200	18.23	0.2833	0.1484	163	0.02	002.06	0
002	Protein, Crude (%)	0045	18.25	0.5000	18.23	0.2833	0.1484	163	0.06	002.06	0
002	Protein, Crude (%)	0208	18.25	0.1000	18.23	0.2833	0.1484	163	0.06	002.06	0
002	Protein, Crude (%)	0510	18.25	0.3000	18.23	0.2833	0.1484	163	0.06	002.06	0
002	Protein, Crude (%)	0610	18.25	0.1000	18.23	0.2833	0.1484	163	0.06	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0589	18.26	0.1100	18.23	0.2833	0.1484	163	0.07	002.06	0
002	Protein, Crude (%)	0123	18.26	0.0100	18.23	0.2833	0.1484	163	0.07	002.06	0
002	Protein, Crude (%)	0004	18.27	0.1800	18.23	0.2833	0.1484	163	0.13	002.06	0
002	Protein, Crude (%)	0144	18.28	0.0700	18.23	0.2833	0.1484	163	0.15	002.06	0
002	Protein, Crude (%)	0953	18.28	0.0400	18.23	0.2833	0.1484	163	0.16	002.06	0
002	Protein, Crude (%)	0242	18.29	0.4000	18.23	0.2833	0.1484	163	0.20	002.06	0
002	Protein, Crude (%)	0142	18.30	0.0700	18.23	0.2833	0.1484	163	0.22	002.06	0
002	Protein, Crude (%)	0610	18.30	0.2000	18.23	0.2833	0.1484	163	0.23	002.01	0
002	Protein, Crude (%)	0033	18.30	0.2000	18.23	0.2833	0.1484	163	0.23	002.06	0
002	Protein, Crude (%)	0160	18.30	0.2000	18.23	0.2833	0.1484	163	0.23	002.06	0
002	Protein, Crude (%)	0345	18.30	0.2000	18.23	0.2833	0.1484	163	0.23	002.06	0
002	Protein, Crude (%)	0674	18.30	0.0000	18.23	0.2833	0.1484	163	0.23	002.05	0
002	Protein, Crude (%)	0047	18.30	0.4000	18.23	0.2833	0.1484	163	0.23	002.06	0
002	Protein, Crude (%)	0687	18.30	0.1000	18.23	0.2833	0.1484	163	0.23	002.06	0
002	Protein, Crude (%)	0353	18.31	0.0300	18.23	0.2833	0.1484	163	0.25	002.06	0
002	Protein, Crude (%)	0019	18.31	0.0600	18.23	0.2833	0.1484	163	0.27	002.06	0
002	Protein, Crude (%)	0009	18.32	0.1500	18.23	0.2833	0.1484	163	0.29	002.06	0
002	Protein, Crude (%)	0164	18.32	0.2500	18.23	0.2833	0.1484	163	0.29	002.06	0
002	Protein, Crude (%)	0027	18.32	0.0970	18.23	0.2833	0.1484	163	0.30	002.06	0
002	Protein, Crude (%)	0226	18.32	0.1200	18.23	0.2833	0.1484	163	0.30	002.06	0
002	Protein, Crude (%)	0726	18.32	0.0200	18.23	0.2833	0.1484	163	0.30	002.06	0
002	Protein, Crude (%)	0037	18.33	0.2900	18.23	0.2833	0.1484	163	0.32	002.06	0
002	Protein, Crude (%)	0651	18.33	0.0215	18.23	0.2833	0.1484	163	0.36	002.00	0
002	Protein, Crude (%)	2073	18.34	0.1500	18.23	0.2833	0.1484	163	0.36	002.05	0
002	Protein, Crude (%)	0175	18.35	0.1000	18.23	0.2833	0.1484	163	0.41	002.06	0
002	Protein, Crude (%)	0574	18.35	0.5000	18.23	0.2833	0.1484	163	0.41	002.06	0
002	Protein, Crude (%)	0857	18.36	0.0500	18.23	0.2833	0.1484	163	0.43	002.06	0
002	Protein, Crude (%)	0001	18.36	0.2600	18.23	0.2833	0.1484	163	0.45	002.06	0
002	Protein, Crude (%)	2290	18.36	0.0800	18.23	0.2833	0.1484	163	0.45	002.11	0
002	Protein, Crude (%)	0675	18.37	0.0100	18.23	0.2833	0.1484	163	0.46	002.06	0
002	Protein, Crude (%)	0018	18.37	0.1800	18.23	0.2833	0.1484	163	0.48	002.06	0
002	Protein, Crude (%)	0083	18.38	0.0700	18.23	0.2833	0.1484	163	0.50	002.06	0
002	Protein, Crude (%)	0872	18.38	0.3700	18.23	0.2833	0.1484	163	0.50	002.06	0
002	Protein, Crude (%)	0571	18.39	0.2900	18.23	0.2833	0.1484	163	0.53	002.06	0
002	Protein, Crude (%)	0298	18.39	0.0200	18.23	0.2833	0.1484	163	0.55	002.06	0
002	Protein, Crude (%)	0034	18.39	0.0590	18.23	0.2833	0.1484	163	0.56	002.06	0
002	Protein, Crude (%)	0910	18.40	0.0000	18.23	0.2833	0.1484	163	0.59	002.06	0
002	Protein, Crude (%)	0918	18.41	0.1400	18.23	0.2833	0.1484	163	0.62	002.06	0
002	Protein, Crude (%)	0309	18.41	0.1515	18.23	0.2833	0.1484	163	0.64	002.06	0
002	Protein, Crude (%)	0968	18.44	0.0480	18.23	0.2833	0.1484	163	0.72	002.06	0
002	Protein, Crude (%)	2213	18.44	0.0400	18.23	0.2833	0.1484	163	0.73	002.06	0
002	Protein, Crude (%)	2081	18.45	0.1900	18.23	0.2833	0.1484	163	0.75	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	2059	18.46	0.2250	18.23	0.2833	0.1484	163	0.81	002.06	0
002	Protein, Crude (%)	0186	18.47	0.1900	18.23	0.2833	0.1484	163	0.82	002.06	0
002	Protein, Crude (%)	0546	18.47	0.0090	18.23	0.2833	0.1484	163	0.84	002.06	0
002	Protein, Crude (%)	0512	18.48	0.1700	18.23	0.2833	0.1484	163	0.85	002.06	0
002	Protein, Crude (%)	0644	18.48	0.1370	18.23	0.2833	0.1484	163	0.86	002.06	0
002	Protein, Crude (%)	0590	18.48	0.2200	18.23	0.2833	0.1484	163	0.87	002.06	0
002	Protein, Crude (%)	0036	18.52	0.0100	18.23	0.2833	0.1484	163	0.99	002.06	0
002	Protein, Crude (%)	0062	18.52	0.0740	18.23	0.2833	0.1484	163	0.99	002.06	0
002	Protein, Crude (%)	2109	18.52	0.3000	18.23	0.2833	0.1484	163	1.01	002.06	0
002	Protein, Crude (%)	2033	18.54	0.2700	18.23	0.2833	0.1484	163	1.06	002.06	0
002	Protein, Crude (%)	0508	18.54	0.2200	18.23	0.2833	0.1484	163	1.07	002.06	0
002	Protein, Crude (%)	0563	18.55	0.3670	18.23	0.2833	0.1484	163	1.13	002.06	0
002	Protein, Crude (%)	0229	18.57	0.0100	18.23	0.2833	0.1484	163	1.17	002.06	0
002	Protein, Crude (%)	0148	18.58	0.0600	18.23	0.2833	0.1484	163	1.22	002.06	0
002	Protein, Crude (%)	2076	18.59	0.0720	18.23	0.2833	0.1484	163	1.27	002.06	0
002	Protein, Crude (%)	0685	18.64	0.0100	18.23	0.2833	0.1484	163	1.42	002.01	0
002	Protein, Crude (%)	0190	18.65	0.0400	18.23	0.2833	0.1484	163	1.47	002.06	0
002	Protein, Crude (%)	0265	18.65	0.1000	18.23	0.2833	0.1484	163	1.47	002.06	0
002	Protein, Crude (%)	0366	18.70	0.6000	18.23	0.2833	0.1484	163	1.65	002.06	0
002	Protein, Crude (%)	0876	18.70	0.0000	18.23	0.2833	0.1484	163	1.65	002.06	0
002	Protein, Crude (%)	0511	18.75	0.4600	18.23	0.2833	0.1484	163	1.82	002.06	0
002	Protein, Crude (%)	0650	18.90	0.0000	18.23	0.2833	0.1484	163	2.35	002.06	0
002	Protein, Crude (%)	2155	18.94	0.1950	18.23	0.2833	0.1484	163	2.49	002.06	0
002	Protein, Crude (%)	0692	18.95	0.3000	18.23	0.2833	0.1484	163	2.53	002.06	0
002	Protein, Crude (%)	0889	18.97	0.5100	18.23	0.2833	0.1484	163	2.58	002.01	0
002	Protein, Crude (%)	0618	19.10	0.1730	18.23	0.2833	0.1484	163	3.04	002.06	0
002	Protein, Crude (%)	0357	19.13	0.2500	18.23	0.2833	0.1484	163	3.15	002.06	0
002	Protein, Crude (%)	0297	19.52	0.2400	18.23	0.2833	0.1484	163	4.54	002.11	0
002	Protein, Crude (%)	0852	19.57	0.0200	18.23	0.2833	0.1484	163	4.72	002.11	0
002	Protein, Crude (%)	0520	19.95	0.1000	18.23	0.2833	0.1484	163	6.06	002.06	0
002	Protein, Crude (%)	2129	20.36	0.1600	18.23	0.2833	0.1484	163	7.51	002.11	0
002	Protein, Crude (%)	0859	20.60	0.2290	18.23	0.2833	0.1484	163	8.35	002.06	0
002	Protein, Crude (%)	0852	21.37	0.1700	18.23	0.2833	0.1484	163	11.05	002.05	0
002	Protein, Crude (%)	0011	21.70	0.0000	18.23	0.2833	0.1484	163	12.24	002.11	0
002	Protein, Crude (%)	0843	19.01	0.8900	18.23	0.2833	0.1484	163	2.72	002.06	1
002	Protein, Crude (%)	2146	12.44	0.0600	18.23	0.2833	0.1484	163	-20.45	002.05	2
002	Protein, Crude (%)	0405	23.91	0.0300	18.23	0.2833	0.1484	163	20.02	002.11	2
003	Fat, Crude (%)	0574	3.035	0.0500	3.780	0.1851	0.0952	129	-4.02	003.14	0
003	Fat, Crude (%)	0843	3.160	0.2400	3.780	0.1851	0.0952	129	-3.35	003.14	0
003	Fat, Crude (%)	0037	3.225	0.1300	3.780	0.1851	0.0952	129	-3.00	003.14	0
003	Fat, Crude (%)	0505	3.270	0.2200	3.780	0.1851	0.0952	129	-2.75	003.14	0
003	Fat, Crude (%)	2129	3.434	0.2190	3.780	0.1851	0.0952	129	-1.87	003.14	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
003	Fat, Crude (%)	0160	3.450	0.5000	3.780	0.1851	0.0952	129	-1.78	003.99	0
003	Fat, Crude (%)	0520	3.470	0.0000	3.780	0.1851	0.0952	129	-1.67	003.14	0
003	Fat, Crude (%)	0619	3.500	0.0600	3.780	0.1851	0.0952	129	-1.51	003.10	0
003	Fat, Crude (%)	0581	3.500	0.1400	3.780	0.1851	0.0952	129	-1.51	003.14	0
003	Fat, Crude (%)	0208	3.505	0.1500	3.780	0.1851	0.0952	129	-1.49	003.13	0
003	Fat, Crude (%)	0034	3.525	0.1700	3.780	0.1851	0.0952	129	-1.38	003.14	0
003	Fat, Crude (%)	0529	3.530	0.2000	3.780	0.1851	0.0952	129	-1.35	003.14	0
003	Fat, Crude (%)	0682	3.540	0.0000	3.780	0.1851	0.0952	129	-1.30	003.06	0
003	Fat, Crude (%)	2193	3.555	0.0500	3.780	0.1851	0.0952	129	-1.21	003.10	0
003	Fat, Crude (%)	0297	3.560	0.3200	3.780	0.1851	0.0952	129	-1.19	003.06	0
003	Fat, Crude (%)	0726	3.570	0.0000	3.780	0.1851	0.0952	129	-1.13	003.14	0
003	Fat, Crude (%)	0009	3.575	0.2900	3.780	0.1851	0.0952	129	-1.11	003.14	0
003	Fat, Crude (%)	0019	3.575	0.0700	3.780	0.1851	0.0952	129	-1.11	003.14	0
003	Fat, Crude (%)	2062	3.591	0.0135	3.780	0.1851	0.0952	129	-1.02	003.12	0
003	Fat, Crude (%)	0590	3.595	0.0700	3.780	0.1851	0.0952	129	-1.00	003.09	0
003	Fat, Crude (%)	2073	3.595	0.0100	3.780	0.1851	0.0952	129	-1.00	003.10	0
003	Fat, Crude (%)	0693	3.603	0.0350	3.780	0.1851	0.0952	129	-0.96	003.10	0
003	Fat, Crude (%)	0074	3.620	0.3400	3.780	0.1851	0.0952	129	-0.86	003.14	0
003	Fat, Crude (%)	0144	3.620	0.3400	3.780	0.1851	0.0952	129	-0.86	003.14	0
003	Fat, Crude (%)	0242	3.625	0.0900	3.780	0.1851	0.0952	129	-0.84	003.10	0
003	Fat, Crude (%)	0263	3.628	0.0086	3.780	0.1851	0.0952	129	-0.82	003.10	0
003	Fat, Crude (%)	0142	3.630	0.0000	3.780	0.1851	0.0952	129	-0.81	003.00	0
003	Fat, Crude (%)	0358	3.630	0.1800	3.780	0.1851	0.0952	129	-0.81	003.14	0
003	Fat, Crude (%)	2188	3.660	0.0400	3.780	0.1851	0.0952	129	-0.65	003.10	0
003	Fat, Crude (%)	0175	3.660	0.0400	3.780	0.1851	0.0952	129	-0.65	003.14	0
003	Fat, Crude (%)	0948	3.660	0.0400	3.780	0.1851	0.0952	129	-0.65	003.14	0
003	Fat, Crude (%)	0036	3.667	0.0216	3.780	0.1851	0.0952	129	-0.61	003.10	0
003	Fat, Crude (%)	0541	3.675	0.0100	3.780	0.1851	0.0952	129	-0.57	003.10	0
003	Fat, Crude (%)	0298	3.680	0.0200	3.780	0.1851	0.0952	129	-0.54	003.14	0
003	Fat, Crude (%)	0049	3.685	0.1500	3.780	0.1851	0.0952	129	-0.51	003.14	0
003	Fat, Crude (%)	2155	3.687	0.0840	3.780	0.1851	0.0952	129	-0.50	003.10	0
003	Fat, Crude (%)	2009	3.692	0.0182	3.780	0.1851	0.0952	129	-0.48	003.10	0
003	Fat, Crude (%)	2076	3.694	0.0310	3.780	0.1851	0.0952	129	-0.46	003.14	0
003	Fat, Crude (%)	0186	3.695	0.0500	3.780	0.1851	0.0952	129	-0.46	003.14	0
003	Fat, Crude (%)	0229	3.699	0.0730	3.780	0.1851	0.0952	129	-0.44	003.14	0
003	Fat, Crude (%)	2199	3.700	0.4000	3.780	0.1851	0.0952	129	-0.43	003.12	0
003	Fat, Crude (%)	0956	3.700	0.2000	3.780	0.1851	0.0952	129	-0.43	003.14	0
003	Fat, Crude (%)	0870	3.701	0.0122	3.780	0.1851	0.0952	129	-0.43	003.10	0
003	Fat, Crude (%)	2181	3.705	0.0500	3.780	0.1851	0.0952	129	-0.40	003.06	0
003	Fat, Crude (%)	0889	3.713	0.0050	3.780	0.1851	0.0952	129	-0.36	003.11	0
003	Fat, Crude (%)	0083	3.715	0.0100	3.780	0.1851	0.0952	129	-0.35	003.14	0
003	Fat, Crude (%)	0511	3.715	0.1100	3.780	0.1851	0.0952	129	-0.35	003.14	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
003	Fat, Crude (%)	0042	3.720	0.3200	3.780	0.1851	0.0952	129	-0.32	003.10	0
003	Fat, Crude (%)	0066	3.730	0.0200	3.780	0.1851	0.0952	129	-0.27	003.14	0
003	Fat, Crude (%)	0939	3.735	0.0900	3.780	0.1851	0.0952	129	-0.24	003.06	0
003	Fat, Crude (%)	0229	3.738	0.0120	3.780	0.1851	0.0952	129	-0.23	003.06	0
003	Fat, Crude (%)	0512	3.740	0.0120	3.780	0.1851	0.0952	129	-0.22	003.09	0
003	Fat, Crude (%)	0674	3.740	0.0000	3.780	0.1851	0.0952	129	-0.22	003.09	0
003	Fat, Crude (%)	0033	3.740	0.0200	3.780	0.1851	0.0952	129	-0.22	003.13	0
003	Fat, Crude (%)	0682	3.740	0.0000	3.780	0.1851	0.0952	129	-0.22	003.14	0
003	Fat, Crude (%)	0646	3.745	0.1100	3.780	0.1851	0.0952	129	-0.19	003.13	0
003	Fat, Crude (%)	0546	3.745	0.0500	3.780	0.1851	0.0952	129	-0.19	003.99	0
003	Fat, Crude (%)	0689	3.750	0.1000	3.780	0.1851	0.0952	129	-0.16	003.06	0
003	Fat, Crude (%)	0045	3.750	0.0800	3.780	0.1851	0.0952	129	-0.16	003.10	0
003	Fat, Crude (%)	0689	3.750	0.1000	3.780	0.1851	0.0952	129	-0.16	003.10	0
003	Fat, Crude (%)	2196	3.750	0.0000	3.780	0.1851	0.0952	129	-0.16	003.10	0
003	Fat, Crude (%)	0098	3.755	0.0500	3.780	0.1851	0.0952	129	-0.13	003.13	0
003	Fat, Crude (%)	0881	3.758	0.1510	3.780	0.1851	0.0952	129	-0.12	003.10	0
003	Fat, Crude (%)	0687	3.760	0.1200	3.780	0.1851	0.0952	129	-0.11	003.06	0
003	Fat, Crude (%)	0028	3.760	0.0000	3.780	0.1851	0.0952	129	-0.11	003.13	0
003	Fat, Crude (%)	0619	3.760	0.0000	3.780	0.1851	0.0952	129	-0.11	003.14	0
003	Fat, Crude (%)	0202	3.775	0.2900	3.780	0.1851	0.0952	129	-0.03	003.14	0
003	Fat, Crude (%)	0035	3.780	0.0800	3.780	0.1851	0.0952	129	0.00	003.14	0
003	Fat, Crude (%)	0027	3.785	0.0300	3.780	0.1851	0.0952	129	0.03	003.09	0
003	Fat, Crude (%)	0653	3.790	0.0000	3.780	0.1851	0.0952	129	0.06	003.00	0
003	Fat, Crude (%)	0598	3.790	0.0200	3.780	0.1851	0.0952	129	0.06	003.14	0
003	Fat, Crude (%)	0563	3.791	0.0674	3.780	0.1851	0.0952	129	0.06	003.06	0
003	Fat, Crude (%)	0563	3.791	0.0674	3.780	0.1851	0.0952	129	0.06	003.14	0
003	Fat, Crude (%)	0123	3.795	0.0900	3.780	0.1851	0.0952	129	0.08	003.13	0
003	Fat, Crude (%)	0357	3.800	0.2000	3.780	0.1851	0.0952	129	0.11	003.14	0
003	Fat, Crude (%)	0876	3.800	0.4000	3.780	0.1851	0.0952	129	0.11	003.14	0
003	Fat, Crude (%)	0878	3.800	0.0400	3.780	0.1851	0.0952	129	0.11	003.14	0
003	Fat, Crude (%)	0098	3.805	0.0100	3.780	0.1851	0.0952	129	0.14	003.10	0
003	Fat, Crude (%)	2006	3.810	0.0200	3.780	0.1851	0.0952	129	0.16	003.13	0
003	Fat, Crude (%)	0164	3.810	0.0400	3.780	0.1851	0.0952	129	0.16	003.00	0
003	Fat, Crude (%)	2033	3.820	0.0000	3.780	0.1851	0.0952	129	0.22	003.06	0
003	Fat, Crude (%)	2053	3.830	0.0200	3.780	0.1851	0.0952	129	0.27	003.00	0
003	Fat, Crude (%)	0004	3.840	0.0200	3.780	0.1851	0.0952	129	0.33	003.00	0
003	Fat, Crude (%)	0910	3.840	0.0800	3.780	0.1851	0.0952	129	0.33	003.06	0
003	Fat, Crude (%)	0948	3.840	0.0000	3.780	0.1851	0.0952	129	0.33	003.09	0
003	Fat, Crude (%)	0366	3.850	0.1000	3.780	0.1851	0.0952	129	0.38	003.10	0
003	Fat, Crude (%)	0897	3.850	0.0000	3.780	0.1851	0.0952	129	0.38	003.10	0
003	Fat, Crude (%)	0100	3.855	0.0300	3.780	0.1851	0.0952	129	0.41	003.10	0
003	Fat, Crude (%)	0407	3.859	0.3450	3.780	0.1851	0.0952	129	0.43	003.14	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
003	Fat, Crude (%)	0148	3.860	0.0600	3.780	0.1851	0.0952	129	0.43	003.14	0
003	Fat, Crude (%)	0004	3.880	0.3000	3.780	0.1851	0.0952	129	0.54	003.14	0
003	Fat, Crude (%)	0644	3.883	0.0170	3.780	0.1851	0.0952	129	0.55	003.06	0
003	Fat, Crude (%)	0098	3.885	0.0500	3.780	0.1851	0.0952	129	0.57	003.09	0
003	Fat, Crude (%)	0003	3.885	0.2900	3.780	0.1851	0.0952	129	0.57	003.14	0
003	Fat, Crude (%)	2192	3.890	0.0400	3.780	0.1851	0.0952	129	0.60	003.00	0
003	Fat, Crude (%)	0309	3.900	0.0000	3.780	0.1851	0.0952	129	0.65	003.00	0
003	Fat, Crude (%)	0226	3.900	0.1600	3.780	0.1851	0.0952	129	0.65	003.09	0
003	Fat, Crude (%)	0692	3.900	0.2000	3.780	0.1851	0.0952	129	0.65	003.14	0
003	Fat, Crude (%)	0853	3.900	0.2000	3.780	0.1851	0.0952	129	0.65	003.14	0
003	Fat, Crude (%)	0638	3.925	0.0500	3.780	0.1851	0.0952	129	0.78	003.09	0
003	Fat, Crude (%)	0175	3.930	0.1200	3.780	0.1851	0.0952	129	0.81	003.00	0
003	Fat, Crude (%)	2066	3.935	0.0500	3.780	0.1851	0.0952	129	0.84	003.06	0
003	Fat, Crude (%)	0171	3.935	0.0300	3.780	0.1851	0.0952	129	0.84	003.12	0
003	Fat, Crude (%)	0265	3.935	0.0700	3.780	0.1851	0.0952	129	0.84	003.14	0
003	Fat, Crude (%)	0968	3.936	0.0060	3.780	0.1851	0.0952	129	0.84	003.14	0
003	Fat, Crude (%)	0889	3.940	0.1800	3.780	0.1851	0.0952	129	0.87	003.10	0
003	Fat, Crude (%)	2109	3.950	0.0800	3.780	0.1851	0.0952	129	0.92	003.00	0
003	Fat, Crude (%)	0510	3.950	0.1000	3.780	0.1851	0.0952	129	0.92	003.09	0
003	Fat, Crude (%)	0190	3.955	0.0100	3.780	0.1851	0.0952	129	0.95	003.14	0
003	Fat, Crude (%)	0148	3.960	0.1400	3.780	0.1851	0.0952	129	0.97	003.10	0
003	Fat, Crude (%)	0353	3.990	0.0600	3.780	0.1851	0.0952	129	1.14	003.09	0
003	Fat, Crude (%)	0345	4.000	0.2000	3.780	0.1851	0.0952	129	1.19	003.00	0
003	Fat, Crude (%)	0278	4.000	0.0000	3.780	0.1851	0.0952	129	1.19	003.14	0
003	Fat, Crude (%)	0001	4.003	0.1560	3.780	0.1851	0.0952	129	1.21	003.14	0
003	Fat, Crude (%)	2291	4.015	0.0100	3.780	0.1851	0.0952	129	1.27	003.13	0
003	Fat, Crude (%)	0354	4.045	0.0300	3.780	0.1851	0.0952	129	1.43	003.09	0
003	Fat, Crude (%)	0536	4.055	0.2300	3.780	0.1851	0.0952	129	1.49	003.99	0
003	Fat, Crude (%)	0675	4.065	0.0700	3.780	0.1851	0.0952	129	1.54	003.09	0
003	Fat, Crude (%)	0169	4.075	0.0700	3.780	0.1851	0.0952	129	1.60	003.06	0
003	Fat, Crude (%)	0918	4.075	0.0100	3.780	0.1851	0.0952	129	1.60	003.06	0
003	Fat, Crude (%)	0685	4.085	0.0300	3.780	0.1851	0.0952	129	1.65	003.09	0
003	Fat, Crude (%)	0011	4.100	0.0000	3.780	0.1851	0.0952	129	1.73	003.11	0
003	Fat, Crude (%)	0017	4.300	0.2000	3.780	0.1851	0.0952	129	2.81	003.00	0
003	Fat, Crude (%)	0062	4.311	0.0730	3.780	0.1851	0.0952	129	2.87	003.10	0
003	Fat, Crude (%)	0852	4.320	0.0000	3.780	0.1851	0.0952	129	2.92	003.11	0
003	Fat, Crude (%)	0941	4.505	0.0900	3.780	0.1851	0.0952	129	3.92	003.99	0
003	Fat, Crude (%)	0405	4.595	0.0100	3.780	0.1851	0.0952	129	4.41	003.11	0
003	Fat, Crude (%)	0618	4.768	0.2920	3.780	0.1851	0.0952	129	5.34	003.10	0
003	Fat, Crude (%)	0297	4.905	0.1500	3.780	0.1851	0.0952	129	6.08	003.11	0
003	Fat, Crude (%)	2304	10.42	2.120	3.780	0.1851	0.0952	129	35.88	003.99	2
004	Fiber, Crude (%)	2199	1.650	0.5000	3.505	0.6193	0.2022	103	-2.99	004.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
004	Fiber, Crude (%)	0265	2.355	0.5100	3.505	0.6193	0.2022	103	-1.86	004.07	0
004	Fiber, Crude (%)	0144	2.425	0.2900	3.505	0.6193	0.2022	103	-1.74	004.07	0
004	Fiber, Crude (%)	0510	2.450	0.1000	3.505	0.6193	0.2022	103	-1.70	004.07	0
004	Fiber, Crude (%)	0405	2.565	0.0500	3.505	0.6193	0.2022	103	-1.52	004.11	0
004	Fiber, Crude (%)	0968	2.697	0.1620	3.505	0.6193	0.2022	103	-1.30	004.07	0
004	Fiber, Crude (%)	2109	2.800	0.0400	3.505	0.6193	0.2022	103	-1.14	004.07	0
004	Fiber, Crude (%)	0610	2.845	0.0700	3.505	0.6193	0.2022	103	-1.07	004.99	0
004	Fiber, Crude (%)	0520	2.850	0.1000	3.505	0.6193	0.2022	103	-1.06	004.07	0
004	Fiber, Crude (%)	0619	2.860	0.1400	3.505	0.6193	0.2022	103	-1.04	004.07	0
004	Fiber, Crude (%)	0910	2.900	0.0000	3.505	0.6193	0.2022	103	-0.98	004.07	0
004	Fiber, Crude (%)	2181	2.905	0.0300	3.505	0.6193	0.2022	103	-0.97	004.07	0
004	Fiber, Crude (%)	0354	2.910	0.0400	3.505	0.6193	0.2022	103	-0.96	004.00	0
004	Fiber, Crude (%)	0921	2.940	0.1000	3.505	0.6193	0.2022	103	-0.91	004.03	0
004	Fiber, Crude (%)	0956	2.950	0.3000	3.505	0.6193	0.2022	103	-0.90	004.07	0
004	Fiber, Crude (%)	0674	2.970	0.0000	3.505	0.6193	0.2022	103	-0.86	004.06	0
004	Fiber, Crude (%)	0123	2.980	0.0000	3.505	0.6193	0.2022	103	-0.85	004.07	0
004	Fiber, Crude (%)	0353	3.000	0.0200	3.505	0.6193	0.2022	103	-0.81	004.03	0
004	Fiber, Crude (%)	0098	3.000	0.1000	3.505	0.6193	0.2022	103	-0.81	004.06	0
004	Fiber, Crude (%)	0689	3.000	0.0000	3.505	0.6193	0.2022	103	-0.81	004.06	0
004	Fiber, Crude (%)	0675	3.005	0.2300	3.505	0.6193	0.2022	103	-0.81	004.07	0
004	Fiber, Crude (%)	0175	3.010	0.0200	3.505	0.6193	0.2022	103	-0.80	004.00	0
004	Fiber, Crude (%)	0708	3.035	0.2100	3.505	0.6193	0.2022	103	-0.76	004.07	0
004	Fiber, Crude (%)	2009	3.078	0.0189	3.505	0.6193	0.2022	103	-0.69	004.07	0
004	Fiber, Crude (%)	0682	3.100	0.0000	3.505	0.6193	0.2022	103	-0.65	004.07	0
004	Fiber, Crude (%)	0726	3.130	0.0000	3.505	0.6193	0.2022	103	-0.60	004.07	0
004	Fiber, Crude (%)	0646	3.155	0.0900	3.505	0.6193	0.2022	103	-0.56	004.07	0
004	Fiber, Crude (%)	0354	3.185	0.0300	3.505	0.6193	0.2022	103	-0.52	004.07	0
004	Fiber, Crude (%)	0870	3.186	0.2431	3.505	0.6193	0.2022	103	-0.51	004.07	0
004	Fiber, Crude (%)	0028	3.200	1.200	3.505	0.6193	0.2022	103	-0.49	004.07	0
004	Fiber, Crude (%)	2188	3.200	0.0000	3.505	0.6193	0.2022	103	-0.49	004.07	0
004	Fiber, Crude (%)	0512	3.215	0.0640	3.505	0.6193	0.2022	103	-0.47	004.06	0
004	Fiber, Crude (%)	2066	3.230	0.3600	3.505	0.6193	0.2022	103	-0.44	004.00	0
004	Fiber, Crude (%)	0881	3.230	0.2560	3.505	0.6193	0.2022	103	-0.44	004.06	0
004	Fiber, Crude (%)	0083	3.235	0.1500	3.505	0.6193	0.2022	103	-0.44	004.07	0
004	Fiber, Crude (%)	0590	3.240	0.6200	3.505	0.6193	0.2022	103	-0.43	004.07	0
004	Fiber, Crude (%)	0164	3.250	0.1000	3.505	0.6193	0.2022	103	-0.41	004.00	0
004	Fiber, Crude (%)	0345	3.250	0.2000	3.505	0.6193	0.2022	103	-0.41	004.00	0
004	Fiber, Crude (%)	0190	3.250	0.1000	3.505	0.6193	0.2022	103	-0.41	004.07	0
004	Fiber, Crude (%)	0171	3.255	0.1300	3.505	0.6193	0.2022	103	-0.40	004.00	0
004	Fiber, Crude (%)	2062	3.272	0.0067	3.505	0.6193	0.2022	103	-0.38	004.06	0
004	Fiber, Crude (%)	2053	3.285	0.0300	3.505	0.6193	0.2022	103	-0.35	004.07	0
004	Fiber, Crude (%)	0941	3.290	0.0200	3.505	0.6193	0.2022	103	-0.35	004.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
004	Fiber, Crude (%)	2196	3.290	0.0000	3.505	0.6193	0.2022	103	-0.35	004.07	0
004	Fiber, Crude (%)	0918	3.295	0.0500	3.505	0.6193	0.2022	103	-0.34	004.06	0
004	Fiber, Crude (%)	0638	3.300	0.2000	3.505	0.6193	0.2022	103	-0.33	004.06	0
004	Fiber, Crude (%)	0948	3.330	0.0200	3.505	0.6193	0.2022	103	-0.28	004.06	0
004	Fiber, Crude (%)	2193	3.345	0.0900	3.505	0.6193	0.2022	103	-0.26	004.07	0
004	Fiber, Crude (%)	0160	3.350	0.9000	3.505	0.6193	0.2022	103	-0.25	004.07	0
004	Fiber, Crude (%)	0027	3.362	0.0750	3.505	0.6193	0.2022	103	-0.23	004.06	0
004	Fiber, Crude (%)	0563	3.369	0.2962	3.505	0.6193	0.2022	103	-0.22	004.07	0
004	Fiber, Crude (%)	0848	3.370	0.0000	3.505	0.6193	0.2022	103	-0.22	004.07	0
004	Fiber, Crude (%)	0675	3.380	0.1600	3.505	0.6193	0.2022	103	-0.20	004.06	0
004	Fiber, Crude (%)	0644	3.382	0.1420	3.505	0.6193	0.2022	103	-0.20	004.07	0
004	Fiber, Crude (%)	0968	3.396	0.1250	3.505	0.6193	0.2022	103	-0.18	004.06	0
004	Fiber, Crude (%)	0017	3.400	0.0000	3.505	0.6193	0.2022	103	-0.17	004.00	0
004	Fiber, Crude (%)	0309	3.400	0.2000	3.505	0.6193	0.2022	103	-0.17	004.00	0
004	Fiber, Crude (%)	0878	3.440	0.0400	3.505	0.6193	0.2022	103	-0.10	004.06	0
004	Fiber, Crude (%)	0366	3.450	0.7000	3.505	0.6193	0.2022	103	-0.09	004.01	0
004	Fiber, Crude (%)	0208	3.490	0.2800	3.505	0.6193	0.2022	103	-0.02	004.00	0
004	Fiber, Crude (%)	0692	3.500	0.4000	3.505	0.6193	0.2022	103	-0.01	004.07	0
004	Fiber, Crude (%)	2259	3.506	0.0130	3.505	0.6193	0.2022	103	0.00	004.07	0
004	Fiber, Crude (%)	0529	3.515	0.0500	3.505	0.6193	0.2022	103	0.02	004.07	0
004	Fiber, Crude (%)	2304	3.515	0.0900	3.505	0.6193	0.2022	103	0.02	004.07	0
004	Fiber, Crude (%)	2006	3.530	0.1800	3.505	0.6193	0.2022	103	0.04	004.06	0
004	Fiber, Crude (%)	2290	3.550	0.1000	3.505	0.6193	0.2022	103	0.07	004.11	0
004	Fiber, Crude (%)	0100	3.600	0.5200	3.505	0.6193	0.2022	103	0.15	004.07	0
004	Fiber, Crude (%)	0034	3.615	0.0100	3.505	0.6193	0.2022	103	0.18	004.07	0
004	Fiber, Crude (%)	0004	3.630	0.4400	3.505	0.6193	0.2022	103	0.20	004.07	0
004	Fiber, Crude (%)	0948	3.655	0.1100	3.505	0.6193	0.2022	103	0.24	004.07	0
004	Fiber, Crude (%)	0019	3.690	0.1000	3.505	0.6193	0.2022	103	0.30	004.07	0
004	Fiber, Crude (%)	2073	3.700	0.0000	3.505	0.6193	0.2022	103	0.32	004.00	0
004	Fiber, Crude (%)	0643	3.735	0.2700	3.505	0.6193	0.2022	103	0.37	004.07	0
004	Fiber, Crude (%)	0263	3.757	0.0106	3.505	0.6193	0.2022	103	0.41	004.07	0
004	Fiber, Crude (%)	0298	3.795	0.0300	3.505	0.6193	0.2022	103	0.47	004.07	0
004	Fiber, Crude (%)	0098	3.820	0.6600	3.505	0.6193	0.2022	103	0.51	004.07	0
004	Fiber, Crude (%)	0074	3.910	0.0200	3.505	0.6193	0.2022	103	0.65	004.07	0
004	Fiber, Crude (%)	0897	3.950	0.0000	3.505	0.6193	0.2022	103	0.72	004.06	0
004	Fiber, Crude (%)	0407	3.965	0.2263	3.505	0.6193	0.2022	103	0.74	004.07	0
004	Fiber, Crude (%)	0011	4.000	0.0000	3.505	0.6193	0.2022	103	0.80	004.11	0
004	Fiber, Crude (%)	0889	4.090	0.0600	3.505	0.6193	0.2022	103	0.95	004.11	0
004	Fiber, Crude (%)	2033	4.100	0.0000	3.505	0.6193	0.2022	103	0.96	004.07	0
004	Fiber, Crude (%)	0852	4.120	0.0400	3.505	0.6193	0.2022	103	0.99	004.11	0
004	Fiber, Crude (%)	0693	4.173	0.0720	3.505	0.6193	0.2022	103	1.08	004.07	0
004	Fiber, Crude (%)	0045	4.235	0.6500	3.505	0.6193	0.2022	103	1.18	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
004	Fiber, Crude (%)	0169	4.250	0.3000	3.505	0.6193	0.2022	103	1.20	004.00	0
004	Fiber, Crude (%)	0009	4.255	0.6700	3.505	0.6193	0.2022	103	1.21	004.07	0
004	Fiber, Crude (%)	0226	4.300	0.0400	3.505	0.6193	0.2022	103	1.28	004.00	0
004	Fiber, Crude (%)	2192	4.360	0.7400	3.505	0.6193	0.2022	103	1.38	004.03	0
004	Fiber, Crude (%)	0876	4.400	0.4000	3.505	0.6193	0.2022	103	1.45	004.07	0
004	Fiber, Crude (%)	0278	4.500	0.0000	3.505	0.6193	0.2022	103	1.61	004.07	0
004	Fiber, Crude (%)	0202	4.535	0.9300	3.505	0.6193	0.2022	103	1.66	004.07	0
004	Fiber, Crude (%)	0229	4.680	0.0200	3.505	0.6193	0.2022	103	1.90	004.07	0
004	Fiber, Crude (%)	2076	4.708	0.2070	3.505	0.6193	0.2022	103	1.94	004.00	0
004	Fiber, Crude (%)	0505	4.725	0.1300	3.505	0.6193	0.2022	103	1.97	004.07	0
004	Fiber, Crude (%)	0939	4.750	0.3400	3.505	0.6193	0.2022	103	2.01	004.07	0
004	Fiber, Crude (%)	0598	4.930	0.1600	3.505	0.6193	0.2022	103	2.30	004.07	0
004	Fiber, Crude (%)	0626	4.995	0.0700	3.505	0.6193	0.2022	103	2.41	004.07	0
004	Fiber, Crude (%)	0581	5.100	0.2000	3.505	0.6193	0.2022	103	2.58	004.07	0
004	Fiber, Crude (%)	2129	5.514	1.217	3.505	0.6193	0.2022	103	3.24	004.11	0
004	Fiber, Crude (%)	0536	5.535	0.9900	3.505	0.6193	0.2022	103	3.28	004.07	0
004	Fiber, Crude (%)	0511	5.600	0.4000	3.505	0.6193	0.2022	103	3.38	004.07	0
004	Fiber, Crude (%)	2146	11.26	0.0500	3.505	0.6193	0.2022	103	12.51	004.99	0
004	Fiber, Crude (%)	0003	2.470	2.260	3.505	0.6193	0.2022	103	-1.67	004.07	1
004	Fiber, Crude (%)	2129	26.73	0.2190	3.505	0.6193	0.2022	103	37.51	004.99	2
005	Ash (%)	0297	8.110	0.3000	12.88	0.6409	0.1397	129	-7.44	005.11	0
005	Ash (%)	0889	8.758	0.1450	12.88	0.6409	0.1397	129	-6.42	005.11	0
005	Ash (%)	2081	9.175	0.1500	12.88	0.6409	0.1397	129	-5.77	005.99	0
005	Ash (%)	0019	9.600	0.0000	12.88	0.6409	0.1397	129	-5.11	005.00	0
005	Ash (%)	0169	9.615	0.2300	12.88	0.6409	0.1397	129	-5.09	005.00	0
005	Ash (%)	0062	9.635	0.0380	12.88	0.6409	0.1397	129	-5.06	005.00	0
005	Ash (%)	0278	9.785	0.2300	12.88	0.6409	0.1397	129	-4.82	005.00	0
005	Ash (%)	0852	10.25	0.0300	12.88	0.6409	0.1397	129	-4.10	005.11	0
005	Ash (%)	0144	10.54	0.0100	12.88	0.6409	0.1397	129	-3.65	005.00	0
005	Ash (%)	0049	10.61	0.2000	12.88	0.6409	0.1397	129	-3.53	005.00	0
005	Ash (%)	0650	10.65	0.1000	12.88	0.6409	0.1397	129	-3.47	005.00	0
005	Ash (%)	0539	10.70	0.7200	12.88	0.6409	0.1397	129	-3.39	005.00	0
005	Ash (%)	0675	10.94	0.4900	12.88	0.6409	0.1397	129	-3.03	005.00	0
005	Ash (%)	0357	11.00	0.2000	12.88	0.6409	0.1397	129	-2.93	005.00	0
005	Ash (%)	0208	11.05	0.3000	12.88	0.6409	0.1397	129	-2.85	005.00	0
005	Ash (%)	2033	11.40	0.0000	12.88	0.6409	0.1397	129	-2.30	005.00	0
005	Ash (%)	0853	11.75	0.1000	12.88	0.6409	0.1397	129	-1.76	005.00	0
005	Ash (%)	0918	11.78	0.1900	12.88	0.6409	0.1397	129	-1.72	005.00	0
005	Ash (%)	0511	11.83	0.4200	12.88	0.6409	0.1397	129	-1.63	005.00	0
005	Ash (%)	0160	11.95	0.7000	12.88	0.6409	0.1397	129	-1.44	005.00	0
005	Ash (%)	0003	11.95	0.1000	12.88	0.6409	0.1397	129	-1.44	005.05	0
005	Ash (%)	0598	12.05	0.0300	12.88	0.6409	0.1397	129	-1.30	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
005	Ash (%)	2073	12.07	0.0200	12.88	0.6409	0.1397	129	-1.26	005.00	0
005	Ash (%)	2066	12.12	0.0600	12.88	0.6409	0.1397	129	-1.18	005.00	0
005	Ash (%)	0964	12.16	0.0470	12.88	0.6409	0.1397	129	-1.11	005.00	0
005	Ash (%)	0876	12.25	0.5000	12.88	0.6409	0.1397	129	-0.98	005.00	0
005	Ash (%)	0693	12.29	0.1080	12.88	0.6409	0.1397	129	-0.92	005.00	0
005	Ash (%)	0574	12.30	0.6000	12.88	0.6409	0.1397	129	-0.90	005.00	0
005	Ash (%)	0622	12.35	0.4024	12.88	0.6409	0.1397	129	-0.81	005.00	0
005	Ash (%)	0164	12.47	0.0400	12.88	0.6409	0.1397	129	-0.63	005.00	0
005	Ash (%)	0541	12.49	0.4200	12.88	0.6409	0.1397	129	-0.60	005.00	0
005	Ash (%)	0674	12.50	0.0000	12.88	0.6409	0.1397	129	-0.59	005.00	0
005	Ash (%)	0242	12.51	0.7300	12.88	0.6409	0.1397	129	-0.58	005.00	0
005	Ash (%)	0066	12.51	0.1700	12.88	0.6409	0.1397	129	-0.58	005.05	0
005	Ash (%)	0004	12.52	0.0900	12.88	0.6409	0.1397	129	-0.56	005.00	0
005	Ash (%)	2053	12.55	0.0600	12.88	0.6409	0.1397	129	-0.51	005.00	0
005	Ash (%)	0651	12.55	0.3705	12.88	0.6409	0.1397	129	-0.50	005.00	0
005	Ash (%)	0953	12.58	0.2300	12.88	0.6409	0.1397	129	-0.47	005.00	0
005	Ash (%)	0682	12.60	0.0000	12.88	0.6409	0.1397	129	-0.43	005.00	0
005	Ash (%)	0881	12.62	0.3640	12.88	0.6409	0.1397	129	-0.39	005.00	0
005	Ash (%)	0358	12.63	0.3400	12.88	0.6409	0.1397	129	-0.38	005.00	0
005	Ash (%)	0015	12.65	0.3700	12.88	0.6409	0.1397	129	-0.36	005.00	0
005	Ash (%)	0027	12.67	0.3170	12.88	0.6409	0.1397	129	-0.32	005.00	0
005	Ash (%)	0148	12.70	0.4000	12.88	0.6409	0.1397	129	-0.27	005.00	0
005	Ash (%)	0186	12.70	0.1000	12.88	0.6409	0.1397	129	-0.27	005.00	0
005	Ash (%)	0366	12.70	0.2000	12.88	0.6409	0.1397	129	-0.27	005.00	0
005	Ash (%)	0646	12.71	0.0300	12.88	0.6409	0.1397	129	-0.27	005.00	0
005	Ash (%)	0638	12.72	0.1600	12.88	0.6409	0.1397	129	-0.24	005.05	0
005	Ash (%)	0175	12.75	0.0100	12.88	0.6409	0.1397	129	-0.20	005.00	0
005	Ash (%)	0083	12.79	0.0700	12.88	0.6409	0.1397	129	-0.14	005.00	0
005	Ash (%)	0870	12.80	0.0078	12.88	0.6409	0.1397	129	-0.11	005.00	0
005	Ash (%)	0142	12.82	0.0500	12.88	0.6409	0.1397	129	-0.09	005.00	0
005	Ash (%)	2150	12.84	0.2600	12.88	0.6409	0.1397	129	-0.06	005.00	0
005	Ash (%)	0726	12.86	0.0000	12.88	0.6409	0.1397	129	-0.02	005.05	0
005	Ash (%)	0510	12.87	0.0700	12.88	0.6409	0.1397	129	-0.02	005.00	0
005	Ash (%)	2304	12.90	0.1300	12.88	0.6409	0.1397	129	0.03	005.03	0
005	Ash (%)	2199	12.90	0.0000	12.88	0.6409	0.1397	129	0.04	005.00	0
005	Ash (%)	2076	12.99	0.0245	12.88	0.6409	0.1397	129	0.18	005.00	0
005	Ash (%)	0589	12.99	0.1600	12.88	0.6409	0.1397	129	0.18	005.00	0
005	Ash (%)	0618	13.00	0.1879	12.88	0.6409	0.1397	129	0.19	005.00	0
005	Ash (%)	2006	13.00	0.1000	12.88	0.6409	0.1397	129	0.19	005.00	0
005	Ash (%)	0229	13.02	0.3500	12.88	0.6409	0.1397	129	0.22	005.00	0
005	Ash (%)	0171	13.03	0.0300	12.88	0.6409	0.1397	129	0.23	005.00	0
005	Ash (%)	2196	13.03	0.0100	12.88	0.6409	0.1397	129	0.23	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
005	Ash (%)	0123	13.05	0.1000	12.88	0.6409	0.1397	129	0.27	005.00	0
005	Ash (%)	0848	13.08	0.0700	12.88	0.6409	0.1397	129	0.31	005.00	0
005	Ash (%)	0407	13.08	0.0034	12.88	0.6409	0.1397	129	0.31	005.00	0
005	Ash (%)	2291	13.08	0.0400	12.88	0.6409	0.1397	129	0.32	005.05	0
005	Ash (%)	0546	13.09	0.0500	12.88	0.6409	0.1397	129	0.33	005.99	0
005	Ash (%)	0035	13.12	0.0000	12.88	0.6409	0.1397	129	0.38	005.00	0
005	Ash (%)	0298	13.12	0.0200	12.88	0.6409	0.1397	129	0.38	005.00	0
005	Ash (%)	0265	13.15	0.1000	12.88	0.6409	0.1397	129	0.43	005.00	0
005	Ash (%)	0001	13.15	0.1090	12.88	0.6409	0.1397	129	0.43	005.00	0
005	Ash (%)	0353	13.16	0.0500	12.88	0.6409	0.1397	129	0.44	005.00	0
005	Ash (%)	2302	13.16	0.1100	12.88	0.6409	0.1397	129	0.44	005.00	0
005	Ash (%)	0345	13.18	0.0500	12.88	0.6409	0.1397	129	0.47	005.00	0
005	Ash (%)	0644	13.18	0.0100	12.88	0.6409	0.1397	129	0.47	005.02	0
005	Ash (%)	0263	13.19	0.0372	12.88	0.6409	0.1397	129	0.50	005.05	0
005	Ash (%)	0017	13.20	0.2000	12.88	0.6409	0.1397	129	0.51	005.00	0
005	Ash (%)	0045	13.20	0.0000	12.88	0.6409	0.1397	129	0.51	005.00	0
005	Ash (%)	0692	13.20	0.2000	12.88	0.6409	0.1397	129	0.51	005.05	0
005	Ash (%)	0354	13.21	0.0600	12.88	0.6409	0.1397	129	0.52	005.00	0
005	Ash (%)	0889	13.21	0.2200	12.88	0.6409	0.1397	129	0.52	005.05	0
005	Ash (%)	0968	13.22	0.0870	12.88	0.6409	0.1397	129	0.53	005.02	0
005	Ash (%)	0520	13.24	0.1100	12.88	0.6409	0.1397	129	0.56	005.00	0
005	Ash (%)	0297	13.25	0.0900	12.88	0.6409	0.1397	129	0.58	005.00	0
005	Ash (%)	0098	13.25	0.0800	12.88	0.6409	0.1397	129	0.58	005.00	0
005	Ash (%)	0689	13.27	0.0300	12.88	0.6409	0.1397	129	0.61	005.05	0
005	Ash (%)	0676	13.27	0.0100	12.88	0.6409	0.1397	129	0.61	005.99	0
005	Ash (%)	0563	13.27	0.0618	12.88	0.6409	0.1397	129	0.61	005.00	0
005	Ash (%)	2259	13.27	0.0190	12.88	0.6409	0.1397	129	0.61	005.00	0
005	Ash (%)	0505	13.28	0.0200	12.88	0.6409	0.1397	129	0.63	005.00	0
005	Ash (%)	2193	13.30	0.0800	12.88	0.6409	0.1397	129	0.66	005.00	0
005	Ash (%)	0100	13.31	0.4900	12.88	0.6409	0.1397	129	0.67	005.00	0
005	Ash (%)	0202	13.32	0.2300	12.88	0.6409	0.1397	129	0.69	005.99	0
005	Ash (%)	2062	13.32	0.0101	12.88	0.6409	0.1397	129	0.69	005.05	0
005	Ash (%)	0878	13.33	0.0600	12.88	0.6409	0.1397	129	0.71	005.05	0
005	Ash (%)	2181	13.34	0.0300	12.88	0.6409	0.1397	129	0.72	005.00	0
005	Ash (%)	2192	13.35	0.0500	12.88	0.6409	0.1397	129	0.73	005.05	0
005	Ash (%)	0047	13.35	0.1000	12.88	0.6409	0.1397	129	0.74	005.00	0
005	Ash (%)	0910	13.35	0.1000	12.88	0.6409	0.1397	129	0.74	005.99	0
005	Ash (%)	0653	13.36	0.0500	12.88	0.6409	0.1397	129	0.75	005.05	0
005	Ash (%)	2109	13.36	0.2100	12.88	0.6409	0.1397	129	0.75	005.05	0
005	Ash (%)	0968	13.37	0.1740	12.88	0.6409	0.1397	129	0.77	005.00	0
005	Ash (%)	0309	13.39	0.0900	12.88	0.6409	0.1397	129	0.79	005.00	0
005	Ash (%)	0939	13.40	0.1500	12.88	0.6409	0.1397	129	0.81	005.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
005	Ash (%)	2290	13.40	0.4000	12.88	0.6409	0.1397	129	0.82	005.11	0
005	Ash (%)	0610	13.40	0.2000	12.88	0.6409	0.1397	129	0.82	005.05	0
005	Ash (%)	0643	13.41	0.0700	12.88	0.6409	0.1397	129	0.83	005.00	0
005	Ash (%)	0139	13.41	0.0500	12.88	0.6409	0.1397	129	0.83	005.00	0
005	Ash (%)	0590	13.44	0.1300	12.88	0.6409	0.1397	129	0.87	005.05	0
005	Ash (%)	0897	13.44	0.0000	12.88	0.6409	0.1397	129	0.88	005.05	0
005	Ash (%)	0941	13.44	0.1400	12.88	0.6409	0.1397	129	0.88	005.05	0
005	Ash (%)	0914	13.45	0.0300	12.88	0.6409	0.1397	129	0.89	005.00	0
005	Ash (%)	0619	13.45	0.1000	12.88	0.6409	0.1397	129	0.90	005.05	0
005	Ash (%)	2009	13.50	0.0064	12.88	0.6409	0.1397	129	0.97	005.05	0
005	Ash (%)	0581	13.52	0.0100	12.88	0.6409	0.1397	129	1.00	005.00	0
005	Ash (%)	0226	13.53	0.2200	12.88	0.6409	0.1397	129	1.02	005.00	0
005	Ash (%)	2188	13.56	0.0200	12.88	0.6409	0.1397	129	1.07	005.00	0
005	Ash (%)	0034	13.57	0.1050	12.88	0.6409	0.1397	129	1.09	005.05	0
005	Ash (%)	2129	13.58	0.0518	12.88	0.6409	0.1397	129	1.10	005.05	0
005	Ash (%)	0948	13.59	0.0200	12.88	0.6409	0.1397	129	1.11	005.00	0
005	Ash (%)	0652	13.60	0.0000	12.88	0.6409	0.1397	129	1.13	005.99	0
005	Ash (%)	0921	13.61	0.1800	12.88	0.6409	0.1397	129	1.15	005.05	0
005	Ash (%)	0685	13.62	0.0100	12.88	0.6409	0.1397	129	1.15	005.05	0
005	Ash (%)	0190	13.63	0.0700	12.88	0.6409	0.1397	129	1.17	005.05	0
005	Ash (%)	0536	13.65	0.0600	12.88	0.6409	0.1397	129	1.21	005.99	0
005	Ash (%)	0852	13.71	0.0700	12.88	0.6409	0.1397	129	1.29	005.00	0
005	Ash (%)	2146	14.02	0.0500	12.88	0.6409	0.1397	129	1.78	005.05	0
005	Ash (%)	0956	12.50	1.000	12.88	0.6409	0.1397	129	-0.59	005.00	1
005	Ash (%)	0529	12.70	1.100	12.88	0.6409	0.1397	129	-0.27	005.00	1
005	Ash (%)	2129	5.088	0.0160	12.88	0.6409	0.1397	129	-12.15	005.11	2
006	Total Sugars (%)	2290	2.300	0.2000	3.583	0.7796	0.3424	5		006.99	0
006	Total Sugars (%)	0910	3.435	0.4900	3.583	0.7796	0.3424	5		006.00	0
006	Total Sugars (%)	0956	3.850	0.3000	3.583	0.7796	0.3424	5		006.99	0
006	Total Sugars (%)	0921	4.085	0.1900	3.583	0.7796	0.3424	5		006.00	0
006	Total Sugars (%)	0407	4.246	0.5320	3.583	0.7796	0.3424	5		006.00	0
008	Fiber, Acid Detergent (%)	0968	3.538	0.0420	4.741	0.5628	0.2349	53	-2.14	008.08	0
008	Fiber, Acid Detergent (%)	0675	3.635	0.0500	4.741	0.5628	0.2349	53	-1.96	008.08	0
008	Fiber, Acid Detergent (%)	0848	3.780	0.0000	4.741	0.5628	0.2349	53	-1.71	008.08	0
008	Fiber, Acid Detergent (%)	0878	3.835	0.2500	4.741	0.5628	0.2349	53	-1.61	008.08	0
008	Fiber, Acid Detergent (%)	0675	3.850	0.0200	4.741	0.5628	0.2349	53	-1.58	008.02	0
008	Fiber, Acid Detergent (%)	0918	3.900	0.0680	4.741	0.5628	0.2349	53	-1.49	008.08	0
008	Fiber, Acid Detergent (%)	0653	3.965	0.0300	4.741	0.5628	0.2349	53	-1.38	008.08	0
008	Fiber, Acid Detergent (%)	0098	3.970	0.0000	4.741	0.5628	0.2349	53	-1.37	008.02	0
008	Fiber, Acid Detergent (%)	0160	4.105	0.2100	4.741	0.5628	0.2349	53	-1.13	008.08	0
008	Fiber, Acid Detergent (%)	0590	4.145	0.6100	4.741	0.5628	0.2349	53	-1.06	008.08	0
008	Fiber, Acid Detergent (%)	0358	4.325	0.1500	4.741	0.5628	0.2349	53	-0.74	008.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
008	Fiber, Acid Detergent (%)	0948	4.330	0.0600	4.741	0.5628	0.2349	53	-0.73	008.08	0
008	Fiber, Acid Detergent (%)	0646	4.470	0.1000	4.741	0.5628	0.2349	53	-0.48	008.08	0
008	Fiber, Acid Detergent (%)	0939	4.470	0.5000	4.741	0.5628	0.2349	53	-0.48	008.08	0
008	Fiber, Acid Detergent (%)	0910	4.490	0.6000	4.741	0.5628	0.2349	53	-0.45	008.08	0
008	Fiber, Acid Detergent (%)	0353	4.495	0.1100	4.741	0.5628	0.2349	53	-0.44	008.02	0
008	Fiber, Acid Detergent (%)	0948	4.495	0.3500	4.741	0.5628	0.2349	53	-0.44	008.02	0
008	Fiber, Acid Detergent (%)	0354	4.500	0.1200	4.741	0.5628	0.2349	53	-0.43	008.08	0
008	Fiber, Acid Detergent (%)	0004	4.510	0.2400	4.741	0.5628	0.2349	53	-0.41	008.08	0
008	Fiber, Acid Detergent (%)	0278	4.600	0.2000	4.741	0.5628	0.2349	53	-0.25	008.08	0
008	Fiber, Acid Detergent (%)	2129	4.628	1.206	4.741	0.5628	0.2349	53	-0.20	008.08	0
008	Fiber, Acid Detergent (%)	0536	4.640	0.4600	4.741	0.5628	0.2349	53	-0.18	008.08	0
008	Fiber, Acid Detergent (%)	0049	4.680	0.2400	4.741	0.5628	0.2349	53	-0.11	008.08	0
008	Fiber, Acid Detergent (%)	0309	4.684	0.1427	4.741	0.5628	0.2349	53	-0.10	008.02	0
008	Fiber, Acid Detergent (%)	0345	4.750	0.3000	4.741	0.5628	0.2349	53	0.02	008.02	0
008	Fiber, Acid Detergent (%)	0164	4.750	0.1000	4.741	0.5628	0.2349	53	0.02	008.08	0
008	Fiber, Acid Detergent (%)	0263	4.793	0.0366	4.741	0.5628	0.2349	53	0.09	008.08	0
008	Fiber, Acid Detergent (%)	0610	4.810	0.2600	4.741	0.5628	0.2349	53	0.12	008.99	0
008	Fiber, Acid Detergent (%)	0083	4.835	0.2700	4.741	0.5628	0.2349	53	0.17	008.08	0
008	Fiber, Acid Detergent (%)	0148	4.870	0.1000	4.741	0.5628	0.2349	53	0.23	008.08	0
008	Fiber, Acid Detergent (%)	0689	4.900	0.6000	4.741	0.5628	0.2349	53	0.28	008.02	0
008	Fiber, Acid Detergent (%)	0042	4.925	0.5100	4.741	0.5628	0.2349	53	0.33	008.08	0
008	Fiber, Acid Detergent (%)	2196	4.940	0.0000	4.741	0.5628	0.2349	53	0.35	008.08	0
008	Fiber, Acid Detergent (%)	0190	4.960	0.0800	4.741	0.5628	0.2349	53	0.39	008.08	0
008	Fiber, Acid Detergent (%)	2192	4.985	0.4100	4.741	0.5628	0.2349	53	0.43	008.02	0
008	Fiber, Acid Detergent (%)	0357	5.050	0.1000	4.741	0.5628	0.2349	53	0.55	008.08	0
008	Fiber, Acid Detergent (%)	0001	5.065	0.1300	4.741	0.5628	0.2349	53	0.58	008.08	0
008	Fiber, Acid Detergent (%)	0407	5.113	0.1750	4.741	0.5628	0.2349	53	0.66	008.08	0
008	Fiber, Acid Detergent (%)	0032	5.145	0.2900	4.741	0.5628	0.2349	53	0.72	008.08	0
008	Fiber, Acid Detergent (%)	0619	5.165	0.3300	4.741	0.5628	0.2349	53	0.75	008.08	0
008	Fiber, Acid Detergent (%)	2193	5.175	0.1900	4.741	0.5628	0.2349	53	0.77	008.08	0
008	Fiber, Acid Detergent (%)	0037	5.190	0.3400	4.741	0.5628	0.2349	53	0.80	008.08	0
008	Fiber, Acid Detergent (%)	0726	5.205	0.0700	4.741	0.5628	0.2349	53	0.82	008.08	0
008	Fiber, Acid Detergent (%)	0226	5.210	0.1200	4.741	0.5628	0.2349	53	0.83	008.02	0
008	Fiber, Acid Detergent (%)	0148	5.220	0.4200	4.741	0.5628	0.2349	53	0.85	008.02	0
008	Fiber, Acid Detergent (%)	0510	5.250	0.1000	4.741	0.5628	0.2349	53	0.90	008.08	0
008	Fiber, Acid Detergent (%)	0265	5.285	0.1300	4.741	0.5628	0.2349	53	0.97	008.05	0
008	Fiber, Acid Detergent (%)	0045	5.295	0.1300	4.741	0.5628	0.2349	53	0.98	008.08	0
008	Fiber, Acid Detergent (%)	0581	5.350	0.0600	4.741	0.5628	0.2349	53	1.08	008.08	0
008	Fiber, Acid Detergent (%)	2290	5.450	0.3000	4.741	0.5628	0.2349	53	1.26	008.99	0
008	Fiber, Acid Detergent (%)	0870	5.521	0.1623	4.741	0.5628	0.2349	53	1.39	008.08	0
008	Fiber, Acid Detergent (%)	0693	5.843	0.4750	4.741	0.5628	0.2349	53	1.96	008.08	0
008	Fiber, Acid Detergent (%)	0511	7.050	0.5000	4.741	0.5628	0.2349	53	4.10	008.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
008	Fiber, Acid Detergent (%)	0956	3.750	1.300	4.741	0.5628	0.2349	53	-1.76	008.08	1
008	Fiber, Acid Detergent (%)	0297	4.090	1.460	4.741	0.5628	0.2349	53	-1.16	008.08	1
009	Fiber, Neutral Detergent (%)	0910	9.850	0.1000	11.68	1.048	0.3122	53	-1.75	009.07	0
009	Fiber, Neutral Detergent (%)	0675	10.11	0.0600	11.68	1.048	0.3122	53	-1.50	009.09	0
009	Fiber, Neutral Detergent (%)	0358	10.25	0.2000	11.68	1.048	0.3122	53	-1.37	009.09	0
009	Fiber, Neutral Detergent (%)	0675	10.28	0.1100	11.68	1.048	0.3122	53	-1.34	009.07	0
009	Fiber, Neutral Detergent (%)	0598	10.32	0.1000	11.68	1.048	0.3122	53	-1.30	009.09	0
009	Fiber, Neutral Detergent (%)	0263	10.33	0.0520	11.68	1.048	0.3122	53	-1.29	009.09	0
009	Fiber, Neutral Detergent (%)	0357	10.45	0.1000	11.68	1.048	0.3122	53	-1.18	009.09	0
009	Fiber, Neutral Detergent (%)	0878	10.52	0.3400	11.68	1.048	0.3122	53	-1.11	009.09	0
009	Fiber, Neutral Detergent (%)	0510	10.70	0.4000	11.68	1.048	0.3122	53	-0.94	009.09	0
009	Fiber, Neutral Detergent (%)	0881	10.77	0.8560	11.68	1.048	0.3122	53	-0.87	009.09	0
009	Fiber, Neutral Detergent (%)	0049	10.85	0.6300	11.68	1.048	0.3122	53	-0.80	009.09	0
009	Fiber, Neutral Detergent (%)	0968	10.90	0.0060	11.68	1.048	0.3122	53	-0.74	009.09	0
009	Fiber, Neutral Detergent (%)	0939	10.95	0.0000	11.68	1.048	0.3122	53	-0.70	009.09	0
009	Fiber, Neutral Detergent (%)	0870	11.08	0.5923	11.68	1.048	0.3122	53	-0.58	009.09	0
009	Fiber, Neutral Detergent (%)	0160	11.20	0.2000	11.68	1.048	0.3122	53	-0.46	009.09	0
009	Fiber, Neutral Detergent (%)	0581	11.22	0.2600	11.68	1.048	0.3122	53	-0.44	009.09	0
009	Fiber, Neutral Detergent (%)	0353	11.29	0.2200	11.68	1.048	0.3122	53	-0.38	009.07	0
009	Fiber, Neutral Detergent (%)	0164	11.30	0.4000	11.68	1.048	0.3122	53	-0.37	009.09	0
009	Fiber, Neutral Detergent (%)	2193	11.31	0.0200	11.68	1.048	0.3122	53	-0.36	009.09	0
009	Fiber, Neutral Detergent (%)	0619	11.33	0.0400	11.68	1.048	0.3122	53	-0.34	009.09	0
009	Fiber, Neutral Detergent (%)	0345	11.39	0.0300	11.68	1.048	0.3122	53	-0.29	009.07	0
009	Fiber, Neutral Detergent (%)	2129	11.39	0.6590	11.68	1.048	0.3122	53	-0.28	009.09	0
009	Fiber, Neutral Detergent (%)	0848	11.45	0.0000	11.68	1.048	0.3122	53	-0.22	009.09	0
009	Fiber, Neutral Detergent (%)	2196	11.53	0.0000	11.68	1.048	0.3122	53	-0.15	009.09	0
009	Fiber, Neutral Detergent (%)	0098	11.60	0.2000	11.68	1.048	0.3122	53	-0.08	009.07	0
009	Fiber, Neutral Detergent (%)	0354	11.61	0.2500	11.68	1.048	0.3122	53	-0.08	009.09	0
009	Fiber, Neutral Detergent (%)	0037	11.63	0.3900	11.68	1.048	0.3122	53	-0.06	009.09	0
009	Fiber, Neutral Detergent (%)	0948	11.65	0.2200	11.68	1.048	0.3122	53	-0.03	009.09	0
009	Fiber, Neutral Detergent (%)	0646	11.73	0.1100	11.68	1.048	0.3122	53	0.04	009.09	0
009	Fiber, Neutral Detergent (%)	0948	11.73	0.8800	11.68	1.048	0.3122	53	0.04	009.07	0
009	Fiber, Neutral Detergent (%)	0083	11.80	0.2700	11.68	1.048	0.3122	53	0.11	009.09	0
009	Fiber, Neutral Detergent (%)	0278	11.85	0.1000	11.68	1.048	0.3122	53	0.16	009.09	0
009	Fiber, Neutral Detergent (%)	0653	11.88	0.2200	11.68	1.048	0.3122	53	0.19	009.09	0
009	Fiber, Neutral Detergent (%)	0148	11.90	0.6000	11.68	1.048	0.3122	53	0.21	009.09	0
009	Fiber, Neutral Detergent (%)	0918	11.91	0.7702	11.68	1.048	0.3122	53	0.22	009.09	0
009	Fiber, Neutral Detergent (%)	0964	11.94	1.496	11.68	1.048	0.3122	53	0.24	009.07	0
009	Fiber, Neutral Detergent (%)	0638	12.00	0.4000	11.68	1.048	0.3122	53	0.30	009.09	0
009	Fiber, Neutral Detergent (%)	0407	12.14	0.2741	11.68	1.048	0.3122	53	0.44	009.09	0
009	Fiber, Neutral Detergent (%)	0536	12.23	0.3700	11.68	1.048	0.3122	53	0.52	009.09	0
009	Fiber, Neutral Detergent (%)	0726	12.27	0.1300	11.68	1.048	0.3122	53	0.55	009.09	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
009	Fiber, Neutral Detergent (%)	0590	12.32	0.2500	11.68	1.048	0.3122	53	0.60	009.09	0
009	Fiber, Neutral Detergent (%)	0693	12.49	0.4530	11.68	1.048	0.3122	53	0.77	009.09	0
009	Fiber, Neutral Detergent (%)	2192	12.62	0.6700	11.68	1.048	0.3122	53	0.89	009.07	0
009	Fiber, Neutral Detergent (%)	0032	12.63	0.0400	11.68	1.048	0.3122	53	0.90	009.09	0
009	Fiber, Neutral Detergent (%)	0226	12.89	0.1700	11.68	1.048	0.3122	53	1.15	009.07	0
009	Fiber, Neutral Detergent (%)	0309	12.93	0.4806	11.68	1.048	0.3122	53	1.19	009.07	0
009	Fiber, Neutral Detergent (%)	2290	13.05	0.3000	11.68	1.048	0.3122	53	1.30	009.99	0
009	Fiber, Neutral Detergent (%)	0045	13.20	0.6000	11.68	1.048	0.3122	53	1.45	009.09	0
009	Fiber, Neutral Detergent (%)	0610	13.60	0.6000	11.68	1.048	0.3122	53	1.83	009.07	0
009	Fiber, Neutral Detergent (%)	0689	13.70	0.6000	11.68	1.048	0.3122	53	1.92	009.07	0
009	Fiber, Neutral Detergent (%)	0265	13.80	0.2000	11.68	1.048	0.3122	53	2.02	009.09	0
009	Fiber, Neutral Detergent (%)	0190	13.93	0.0300	11.68	1.048	0.3122	53	2.14	009.09	0
009	Fiber, Neutral Detergent (%)	0889	13.96	0.1000	11.68	1.048	0.3122	53	2.17	009.99	0
009	Fiber, Neutral Detergent (%)	0956	10.85	2.500	11.68	1.048	0.3122	53	-0.80	009.09	1
009	Fiber, Neutral Detergent (%)	0297	11.07	2.740	11.68	1.048	0.3122	53	-0.59	009.09	1
010	Moisture (%)	0546	8.620	0.3400	9.563	0.6209	0.0980	19	-1.52	010.99	0
010	Moisture (%)	0889	8.758	0.0150	9.563	0.6209	0.0980	19	-1.30	010.11	0
010	Moisture (%)	0843	8.820	0.1600	9.563	0.6209	0.0980	19	-1.20	010.03	0
010	Moisture (%)	0726	8.945	0.0300	9.563	0.6209	0.0980	19	-1.00	010.99	0
010	Moisture (%)	0939	9.260	0.1800	9.563	0.6209	0.0980	19	-0.49	010.99	0
010	Moisture (%)	0948	9.280	0.0200	9.563	0.6209	0.0980	19	-0.46	010.99	0
010	Moisture (%)	2129	9.344	0.0870	9.563	0.6209	0.0980	19	-0.35	010.99	0
010	Moisture (%)	0164	9.480	0.0800	9.563	0.6209	0.0980	19	-0.13	010.03	0
010	Moisture (%)	0652	9.500	0.2000	9.563	0.6209	0.0980	19	-0.10	010.99	0
010	Moisture (%)	0692	9.570	0.0600	9.563	0.6209	0.0980	19	0.01	010.99	0
010	Moisture (%)	2150	9.665	0.0300	9.563	0.6209	0.0980	19	0.16	010.99	0
010	Moisture (%)	2290	9.680	0.2000	9.563	0.6209	0.0980	19	0.19	010.11	0
010	Moisture (%)	0921	9.775	0.0300	9.563	0.6209	0.0980	19	0.34	010.99	0
010	Moisture (%)	2076	9.810	0.0600	9.563	0.6209	0.0980	19	0.40	010.99	0
010	Moisture (%)	0190	9.935	0.0100	9.563	0.6209	0.0980	19	0.60	010.99	0
010	Moisture (%)	0914	9.945	0.0100	9.563	0.6209	0.0980	19	0.61	010.99	0
010	Moisture (%)	0160	10.32	0.2900	9.563	0.6209	0.0980	19	1.21	010.99	0
010	Moisture (%)	0852	10.75	0.0300	9.563	0.6209	0.0980	19	1.90	010.11	0
010	Moisture (%)	0405	11.14	0.0300	9.563	0.6209	0.0980	19	2.53	010.99	0
010	Moisture (%)	2129	9.915	0.6400	9.563	0.6209	0.0980	19	0.57	010.11	1
011	Loss on Drying (%)	0536	8.505	0.0700	9.993	0.3372	0.0752	65	-4.41	011.01	0
011	Loss on Drying (%)	0574	8.510	0.0400	9.993	0.3372	0.0752	65	-4.40	011.01	0
011	Loss on Drying (%)	0643	8.750	0.1000	9.993	0.3372	0.0752	65	-3.69	011.01	0
011	Loss on Drying (%)	0520	8.800	0.0000	9.993	0.3372	0.0752	65	-3.54	011.01	0
011	Loss on Drying (%)	0144	9.530	0.3400	9.993	0.3372	0.0752	65	-1.37	011.01	0
011	Loss on Drying (%)	0650	9.540	0.1400	9.993	0.3372	0.0752	65	-1.34	011.01	0
011	Loss on Drying (%)	2291	9.545	0.0500	9.993	0.3372	0.0752	65	-1.33	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
011	Loss on Drying (%)	0510	9.550	0.1000	9.993	0.3372	0.0752	65	-1.31	011.01	0
011	Loss on Drying (%)	0563	9.598	0.0199	9.993	0.3372	0.0752	65	-1.17	011.01	0
011	Loss on Drying (%)	0298	9.610	0.0200	9.993	0.3372	0.0752	65	-1.14	011.01	0
011	Loss on Drying (%)	0968	9.682	0.0280	9.993	0.3372	0.0752	65	-0.92	011.01	0
011	Loss on Drying (%)	2303	9.700	0.4060	9.993	0.3372	0.0752	65	-0.87	011.99	0
011	Loss on Drying (%)	0651	9.726	0.1459	9.993	0.3372	0.0752	65	-0.79	011.01	0
011	Loss on Drying (%)	2181	9.740	0.0600	9.993	0.3372	0.0752	65	-0.75	011.01	0
011	Loss on Drying (%)	2199	9.750	0.1000	9.993	0.3372	0.0752	65	-0.72	011.01	0
011	Loss on Drying (%)	0857	9.750	0.3000	9.993	0.3372	0.0752	65	-0.72	011.99	0
011	Loss on Drying (%)	2006	9.800	0.1000	9.993	0.3372	0.0752	65	-0.57	011.01	0
011	Loss on Drying (%)	2033	9.805	0.0100	9.993	0.3372	0.0752	65	-0.56	011.01	0
011	Loss on Drying (%)	0646	9.845	0.0500	9.993	0.3372	0.0752	65	-0.44	011.01	0
011	Loss on Drying (%)	0539	9.850	0.3200	9.993	0.3372	0.0752	65	-0.42	011.01	0
011	Loss on Drying (%)	0175	9.850	0.1000	9.993	0.3372	0.0752	65	-0.42	011.01	0
011	Loss on Drying (%)	0942	9.860	0.0800	9.993	0.3372	0.0752	65	-0.39	011.02	0
011	Loss on Drying (%)	0622	9.876	0.1067	9.993	0.3372	0.0752	65	-0.35	011.01	0
011	Loss on Drying (%)	0859	9.878	0.0440	9.993	0.3372	0.0752	65	-0.34	011.01	0
011	Loss on Drying (%)	0881	9.907	0.1780	9.993	0.3372	0.0752	65	-0.26	011.01	0
011	Loss on Drying (%)	0653	9.915	0.0300	9.993	0.3372	0.0752	65	-0.23	011.01	0
011	Loss on Drying (%)	0953	9.930	0.0600	9.993	0.3372	0.0752	65	-0.19	011.01	0
011	Loss on Drying (%)	0263	9.935	0.0100	9.993	0.3372	0.0752	65	-0.17	011.01	0
011	Loss on Drying (%)	2066	9.940	0.0400	9.993	0.3372	0.0752	65	-0.16	011.02	0
011	Loss on Drying (%)	0598	9.950	0.0200	9.993	0.3372	0.0752	65	-0.13	011.01	0
011	Loss on Drying (%)	2168	9.955	0.1100	9.993	0.3372	0.0752	65	-0.11	011.01	0
011	Loss on Drying (%)	0674	9.960	0.0000	9.993	0.3372	0.0752	65	-0.10	011.01	0
011	Loss on Drying (%)	0685	9.965	0.0100	9.993	0.3372	0.0752	65	-0.08	011.01	0
011	Loss on Drying (%)	0226	9.985	0.1500	9.993	0.3372	0.0752	65	-0.02	011.01	0
011	Loss on Drying (%)	0208	10.00	0.0000	9.993	0.3372	0.0752	65	0.02	011.01	0
011	Loss on Drying (%)	0852	10.03	0.0300	9.993	0.3372	0.0752	65	0.09	011.99	0
011	Loss on Drying (%)	2196	10.03	0.0000	9.993	0.3372	0.0752	65	0.11	011.01	0
011	Loss on Drying (%)	0682	10.06	0.0000	9.993	0.3372	0.0752	65	0.20	011.01	0
011	Loss on Drying (%)	2073	10.07	0.0100	9.993	0.3372	0.0752	65	0.21	011.01	0
011	Loss on Drying (%)	0675	10.08	0.0100	9.993	0.3372	0.0752	65	0.24	011.01	0
011	Loss on Drying (%)	0171	10.09	0.0400	9.993	0.3372	0.0752	65	0.29	011.01	0
011	Loss on Drying (%)	0265	10.10	0.0000	9.993	0.3372	0.0752	65	0.32	011.01	0
011	Loss on Drying (%)	0511	10.13	0.1500	9.993	0.3372	0.0752	65	0.39	011.01	0
011	Loss on Drying (%)	0354	10.14	0.0100	9.993	0.3372	0.0752	65	0.42	011.01	0
011	Loss on Drying (%)	0618	10.14	0.0600	9.993	0.3372	0.0752	65	0.44	011.01	0
011	Loss on Drying (%)	0100	10.15	0.1100	9.993	0.3372	0.0752	65	0.45	011.01	0
011	Loss on Drying (%)	0164	10.15	0.0800	9.993	0.3372	0.0752	65	0.47	011.01	0
011	Loss on Drying (%)	0098	10.17	0.0200	9.993	0.3372	0.0752	65	0.52	011.01	0
011	Loss on Drying (%)	0529	10.17	0.0000	9.993	0.3372	0.0752	65	0.52	011.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
011	Loss on Drying (%)	0242	10.20	0.2000	9.993	0.3372	0.0752	65	0.61	011.01	0
011	Loss on Drying (%)	0358	10.23	0.0000	9.993	0.3372	0.0752	65	0.70	011.01	0
011	Loss on Drying (%)	2193	10.26	0.0200	9.993	0.3372	0.0752	65	0.79	011.01	0
011	Loss on Drying (%)	0066	10.27	0.0100	9.993	0.3372	0.0752	65	0.81	011.01	0
011	Loss on Drying (%)	0309	10.31	0.0500	9.993	0.3372	0.0752	65	0.92	011.01	0
011	Loss on Drying (%)	0123	10.34	0.0000	9.993	0.3372	0.0752	65	1.03	011.01	0
011	Loss on Drying (%)	0229	10.38	0.0050	9.993	0.3372	0.0752	65	1.15	011.01	0
011	Loss on Drying (%)	0589	10.39	0.0200	9.993	0.3372	0.0752	65	1.18	011.01	0
011	Loss on Drying (%)	0870	10.41	0.1630	9.993	0.3372	0.0752	65	1.24	011.01	0
011	Loss on Drying (%)	0541	10.42	0.1600	9.993	0.3372	0.0752	65	1.27	011.01	0
011	Loss on Drying (%)	0202	10.51	0.1400	9.993	0.3372	0.0752	65	1.53	011.01	0
011	Loss on Drying (%)	0407	10.53	0.0128	9.993	0.3372	0.0752	65	1.58	011.01	0
011	Loss on Drying (%)	0843	10.53	0.0200	9.993	0.3372	0.0752	65	1.59	011.01	0
011	Loss on Drying (%)	0004	10.54	0.0500	9.993	0.3372	0.0752	65	1.61	011.01	0
011	Loss on Drying (%)	0062	13.71	0.1700	9.993	0.3372	0.0752	65	11.02	011.01	0
011	Loss on Drying (%)	2188	13.89	0.0100	9.993	0.3372	0.0752	65	11.54	011.01	0
011	Loss on Drying (%)	2304	8.805	0.4900	9.993	0.3372	0.0752	65	-3.52	011.99	1
011	Loss on Drying (%)	0047	10.05	0.5000	9.993	0.3372	0.0752	65	0.17	011.01	1
012	Starch (%)	0910	27.75	0.5000	33.12	1.949	0.3895	37	-2.75	012.01	0
012	Starch (%)	0918	28.23	0.1250	33.12	1.949	0.3895	37	-2.51	012.01	0
012	Starch (%)	0510	29.20	0.2000	33.12	1.949	0.3895	37	-2.01	012.04	0
012	Starch (%)	0407	29.80	0.1333	33.12	1.949	0.3895	37	-1.70	012.03	0
012	Starch (%)	0689	30.95	0.3000	33.12	1.949	0.3895	37	-1.11	012.00	0
012	Starch (%)	0956	31.00	1.800	33.12	1.949	0.3895	37	-1.09	012.01	0
012	Starch (%)	0265	31.05	0.3000	33.12	1.949	0.3895	37	-1.06	012.01	0
012	Starch (%)	0870	31.35	0.5480	33.12	1.949	0.3895	37	-0.91	012.01	0
012	Starch (%)	0098	31.75	0.1000	33.12	1.949	0.3895	37	-0.70	012.03	0
012	Starch (%)	0353	31.89	0.1000	33.12	1.949	0.3895	37	-0.63	012.04	0
012	Starch (%)	0353	31.89	0.1000	33.12	1.949	0.3895	37	-0.63	012.20	0
012	Starch (%)	0354	32.08	0.1900	33.12	1.949	0.3895	37	-0.53	012.00	0
012	Starch (%)	0278	32.15	0.9000	33.12	1.949	0.3895	37	-0.50	012.04	0
012	Starch (%)	0948	32.29	1.010	33.12	1.949	0.3895	37	-0.43	012.04	0
012	Starch (%)	0045	32.65	0.3000	33.12	1.949	0.3895	37	-0.24	012.01	0
012	Starch (%)	0148	33.05	0.1000	33.12	1.949	0.3895	37	-0.03	012.01	0
012	Starch (%)	0693	33.30	1.011	33.12	1.949	0.3895	37	0.10	012.01	0
012	Starch (%)	0297	33.33	0.5600	33.12	1.949	0.3895	37	0.11	012.03	0
012	Starch (%)	2129	33.42	0.5620	33.12	1.949	0.3895	37	0.15	012.04	0
012	Starch (%)	0164	33.60	0.2000	33.12	1.949	0.3895	37	0.25	012.00	0
012	Starch (%)	2196	33.74	0.0000	33.12	1.949	0.3895	37	0.32	012.00	0
012	Starch (%)	0619	33.80	0.2000	33.12	1.949	0.3895	37	0.35	012.00	0
012	Starch (%)	2129	33.92	0.9040	33.12	1.949	0.3895	37	0.41	012.11	0
012	Starch (%)	2181	33.97	0.3300	33.12	1.949	0.3895	37	0.43	012.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
012	Starch (%)	2006	34.08	0.1600	33.12	1.949	0.3895	37	0.49	012.00	0
012	Starch (%)	0610	34.45	0.3000	33.12	1.949	0.3895	37	0.68	012.00	0
012	Starch (%)	0889	34.47	0.0600	33.12	1.949	0.3895	37	0.69	012.03	0
012	Starch (%)	0878	34.50	0.7600	33.12	1.949	0.3895	37	0.71	012.00	0
012	Starch (%)	2009	34.60	0.1332	33.12	1.949	0.3895	37	0.76	012.00	0
012	Starch (%)	2188	34.63	0.0000	33.12	1.949	0.3895	37	0.78	012.00	0
012	Starch (%)	2193	34.63	0.2200	33.12	1.949	0.3895	37	0.78	012.00	0
012	Starch (%)	0910	34.65	1.300	33.12	1.949	0.3895	37	0.79	012.00	0
012	Starch (%)	0889	34.75	0.4950	33.12	1.949	0.3895	37	0.84	012.11	0
012	Starch (%)	2192	34.81	0.3900	33.12	1.949	0.3895	37	0.87	012.00	0
012	Starch (%)	2290	35.75	0.1000	33.12	1.949	0.3895	37	1.35	012.11	0
012	Starch (%)	0852	36.71	0.0100	33.12	1.949	0.3895	37	1.84	012.11	0
012	Starch (%)	0297	37.86	0.0100	33.12	1.949	0.3895	37	2.43	012.11	0
012	Starch (%)	0164	30.35	2.100	33.12	1.949	0.3895	37	-1.42	012.01	1
013	Fat, Acid Pretreat (%)	0618	3.253	0.2463	4.739	0.5301	0.1556	48	-2.80	013.00	0
013	Fat, Acid Pretreat (%)	0036	3.719	0.1520	4.739	0.5301	0.1556	48	-1.92	013.10	0
013	Fat, Acid Pretreat (%)	0581	3.745	0.2100	4.739	0.5301	0.1556	48	-1.88	013.13	0
013	Fat, Acid Pretreat (%)	2073	3.850	0.0200	4.739	0.5301	0.1556	48	-1.68	013.00	0
013	Fat, Acid Pretreat (%)	2146	4.080	0.2000	4.739	0.5301	0.1556	48	-1.24	013.00	0
013	Fat, Acid Pretreat (%)	2290	4.100	0.0000	4.739	0.5301	0.1556	48	-1.21	013.12	0
013	Fat, Acid Pretreat (%)	0407	4.234	0.0389	4.739	0.5301	0.1556	48	-0.95	013.13	0
013	Fat, Acid Pretreat (%)	0689	4.300	0.0000	4.739	0.5301	0.1556	48	-0.83	013.00	0
013	Fat, Acid Pretreat (%)	0083	4.330	0.0400	4.739	0.5301	0.1556	48	-0.77	013.00	0
013	Fat, Acid Pretreat (%)	0650	4.350	0.2200	4.739	0.5301	0.1556	48	-0.73	013.02	0
013	Fat, Acid Pretreat (%)	0853	4.350	0.1000	4.739	0.5301	0.1556	48	-0.73	013.02	0
013	Fat, Acid Pretreat (%)	0638	4.450	0.1000	4.739	0.5301	0.1556	48	-0.55	013.10	0
013	Fat, Acid Pretreat (%)	0541	4.455	0.0500	4.739	0.5301	0.1556	48	-0.54	013.00	0
013	Fat, Acid Pretreat (%)	0843	4.475	0.5500	4.739	0.5301	0.1556	48	-0.50	013.13	0
013	Fat, Acid Pretreat (%)	0242	4.500	0.4600	4.739	0.5301	0.1556	48	-0.45	013.00	0
013	Fat, Acid Pretreat (%)	0910	4.500	0.0200	4.739	0.5301	0.1556	48	-0.45	013.00	0
013	Fat, Acid Pretreat (%)	0353	4.505	0.1100	4.739	0.5301	0.1556	48	-0.44	013.10	0
013	Fat, Acid Pretreat (%)	0309	4.510	0.1112	4.739	0.5301	0.1556	48	-0.43	013.00	0
013	Fat, Acid Pretreat (%)	0939	4.520	0.1400	4.739	0.5301	0.1556	48	-0.41	013.10	0
013	Fat, Acid Pretreat (%)	0610	4.575	0.0300	4.739	0.5301	0.1556	48	-0.31	013.10	0
013	Fat, Acid Pretreat (%)	0652	4.600	0.0000	4.739	0.5301	0.1556	48	-0.26	013.00	0
013	Fat, Acid Pretreat (%)	2196	4.600	0.0000	4.739	0.5301	0.1556	48	-0.26	013.10	0
013	Fat, Acid Pretreat (%)	0297	4.625	0.4500	4.739	0.5301	0.1556	48	-0.22	013.00	0
013	Fat, Acid Pretreat (%)	0968	4.661	0.0050	4.739	0.5301	0.1556	48	-0.15	013.13	0
013	Fat, Acid Pretreat (%)	0100	4.700	0.0600	4.739	0.5301	0.1556	48	-0.07	013.02	0
013	Fat, Acid Pretreat (%)	0921	4.720	0.2400	4.739	0.5301	0.1556	48	-0.04	013.02	0
013	Fat, Acid Pretreat (%)	0948	4.745	0.0500	4.739	0.5301	0.1556	48	0.01	013.00	0
013	Fat, Acid Pretreat (%)	0354	4.750	0.2400	4.739	0.5301	0.1556	48	0.02	013.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
013	Fat, Acid Pretreat (%)	0098	4.785	0.1500	4.739	0.5301	0.1556	48	0.09	013.02	0
013	Fat, Acid Pretreat (%)	0208	4.870	0.1400	4.739	0.5301	0.1556	48	0.25	013.02	0
013	Fat, Acid Pretreat (%)	0229	4.915	0.1700	4.739	0.5301	0.1556	48	0.33	013.00	0
013	Fat, Acid Pretreat (%)	0148	4.940	0.0800	4.739	0.5301	0.1556	48	0.38	013.02	0
013	Fat, Acid Pretreat (%)	0675	5.015	0.1900	4.739	0.5301	0.1556	48	0.52	013.02	0
013	Fat, Acid Pretreat (%)	0407	5.073	0.0930	4.739	0.5301	0.1556	48	0.63	013.02	0
013	Fat, Acid Pretreat (%)	0265	5.140	0.2200	4.739	0.5301	0.1556	48	0.76	013.13	0
013	Fat, Acid Pretreat (%)	2150	5.145	0.0500	4.739	0.5301	0.1556	48	0.77	013.00	0
013	Fat, Acid Pretreat (%)	0164	5.145	0.0700	4.739	0.5301	0.1556	48	0.77	013.02	0
013	Fat, Acid Pretreat (%)	0870	5.153	0.3599	4.739	0.5301	0.1556	48	0.78	013.02	0
013	Fat, Acid Pretreat (%)	0682	5.160	0.0000	4.739	0.5301	0.1556	48	0.79	013.02	0
013	Fat, Acid Pretreat (%)	0202	5.220	0.2800	4.739	0.5301	0.1556	48	0.91	013.00	0
013	Fat, Acid Pretreat (%)	2076	5.236	0.5112	4.739	0.5301	0.1556	48	0.94	013.00	0
013	Fat, Acid Pretreat (%)	0139	5.250	0.0400	4.739	0.5301	0.1556	48	0.96	013.00	0
013	Fat, Acid Pretreat (%)	0010	5.350	0.3000	4.739	0.5301	0.1556	48	1.15	013.02	0
013	Fat, Acid Pretreat (%)	0914	5.565	0.1100	4.739	0.5301	0.1556	48	1.56	013.02	0
013	Fat, Acid Pretreat (%)	0171	5.590	0.2000	4.739	0.5301	0.1556	48	1.61	013.02	0
013	Fat, Acid Pretreat (%)	2259	5.643	0.1240	4.739	0.5301	0.1556	48	1.70	013.02	0
013	Fat, Acid Pretreat (%)	0643	5.790	0.3000	4.739	0.5301	0.1556	48	1.98	013.02	0
013	Fat, Acid Pretreat (%)	0618	7.104	0.2378	4.739	0.5301	0.1556	48	4.46	013.08	0
015	Aluminum (ppm)	0037	50.20	2.400	209.7	61.51	8.134	15	-2.59	015.42	0
015	Aluminum (ppm)	2129	88.91	1.090	209.7	61.51	8.134	15	-1.96	015.42	0
015	Aluminum (ppm)	0560	120.7	19.60	209.7	61.51	8.134	15	-1.45	015.52	0
015	Aluminum (ppm)	0345	199.0	6.000	209.7	61.51	8.134	15	-0.17	015.43	0
015	Aluminum (ppm)	0049	199.2	8.110	209.7	61.51	8.134	15	-0.17	015.41	0
015	Aluminum (ppm)	0407	202.5	12.14	209.7	61.51	8.134	15	-0.12	015.41	0
015	Aluminum (ppm)	0353	207.0	6.000	209.7	61.51	8.134	15	-0.04	015.43	0
015	Aluminum (ppm)	0169	213.0	2.000	209.7	61.51	8.134	15	0.05	015.43	0
015	Aluminum (ppm)	0407	221.2	3.190	209.7	61.51	8.134	15	0.19	015.43	0
015	Aluminum (ppm)	0297	222.0	6.000	209.7	61.51	8.134	15	0.20	015.43	0
015	Aluminum (ppm)	0510	252.5	1.000	209.7	61.51	8.134	15	0.70	015.43	0
015	Aluminum (ppm)	0964	264.8	21.29	209.7	61.51	8.134	15	0.89	015.43	0
015	Aluminum (ppm)	0520	268.5	7.000	209.7	61.51	8.134	15	0.96	015.41	0
015	Aluminum (ppm)	0164	269.5	23.00	209.7	61.51	8.134	15	0.97	015.41	0
015	Aluminum (ppm)	0171	271.2	3.200	209.7	61.51	8.134	15	1.00	015.41	0
017	Boron (ppm)	0353	9.080	0.1800	10.92	1.169	0.3446	18	-1.57	017.43	0
017	Boron (ppm)	0918	9.520	0.0800	10.92	1.169	0.3446	18	-1.20	017.43	0
017	Boron (ppm)	0037	9.975	0.0760	10.92	1.169	0.3446	18	-0.81	017.42	0
017	Boron (ppm)	0407	10.04	0.1231	10.92	1.169	0.3446	18	-0.75	017.41	0
017	Boron (ppm)	0345	10.10	0.4000	10.92	1.169	0.3446	18	-0.70	017.43	0
017	Boron (ppm)	0407	10.13	0.2011	10.92	1.169	0.3446	18	-0.68	017.43	0
017	Boron (ppm)	0405	10.24	0.1100	10.92	1.169	0.3446	18	-0.59	017.44	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
017	Boron (ppm)	0693	10.55	0.5730	10.92	1.169	0.3446	18	-0.32	017.42	0
017	Boron (ppm)	0560	10.98	0.6900	10.92	1.169	0.3446	18	0.05	017.52	0
017	Boron (ppm)	0510	11.00	0.0000	10.92	1.169	0.3446	18	0.07	017.43	0
017	Boron (ppm)	2129	11.14	0.9000	10.92	1.169	0.3446	18	0.19	017.42	0
017	Boron (ppm)	0045	11.30	1.200	10.92	1.169	0.3446	18	0.33	017.42	0
017	Boron (ppm)	0226	11.58	0.1400	10.92	1.169	0.3446	18	0.56	017.41	0
017	Boron (ppm)	0083	11.65	0.1000	10.92	1.169	0.3446	18	0.62	017.43	0
017	Boron (ppm)	0358	11.68	0.7600	10.92	1.169	0.3446	18	0.65	017.41	0
017	Boron (ppm)	0049	12.17	0.4400	10.92	1.169	0.3446	18	1.07	017.41	0
017	Boron (ppm)	0229	14.05	0.0300	10.92	1.169	0.3446	18	2.67	017.41	0
017	Boron (ppm)	2053	14.50	0.2000	10.92	1.169	0.3446	18	3.06	017.42	0
017	Boron (ppm)	0265	10.95	3.700	10.92	1.169	0.3446	18	0.03	017.42	1
017	Boron (ppm)	0297	0.0000	0.0000	10.92	1.169	0.3446	18		017.43	4
019	Calcium (%)	2146	2.145	0.0300	4.138	0.2184	0.0981	117	-9.13	019.31	0
019	Calcium (%)	2146	2.210	0.1000	4.138	0.2184	0.0981	117	-8.83	019.51	0
019	Calcium (%)	0511	3.190	0.2600	4.138	0.2184	0.0981	117	-4.34	019.41	0
019	Calcium (%)	2062	3.281	0.2298	4.138	0.2184	0.0981	117	-3.92	019.00	0
019	Calcium (%)	0405	3.526	0.0450	4.138	0.2184	0.0981	117	-2.81	019.41	0
019	Calcium (%)	0038	3.688	0.0460	4.138	0.2184	0.0981	117	-2.06	019.43	0
019	Calcium (%)	0563	3.785	0.0364	4.138	0.2184	0.0981	117	-1.62	019.31	0
019	Calcium (%)	0298	3.800	0.0200	4.138	0.2184	0.0981	117	-1.55	019.41	0
019	Calcium (%)	0897	3.810	0.0000	4.138	0.2184	0.0981	117	-1.50	019.00	0
019	Calcium (%)	0036	3.825	0.0686	4.138	0.2184	0.0981	117	-1.44	019.43	0
019	Calcium (%)	0123	3.835	0.0500	4.138	0.2184	0.0981	117	-1.39	019.41	0
019	Calcium (%)	0852	3.850	0.1000	4.138	0.2184	0.0981	117	-1.32	019.08	0
019	Calcium (%)	0910	3.865	0.4700	4.138	0.2184	0.0981	117	-1.25	019.41	0
019	Calcium (%)	0171	3.875	0.0100	4.138	0.2184	0.0981	117	-1.21	019.41	0
019	Calcium (%)	0939	3.880	0.0800	4.138	0.2184	0.0981	117	-1.18	019.31	0
019	Calcium (%)	0004	3.890	0.4600	4.138	0.2184	0.0981	117	-1.14	019.41	0
019	Calcium (%)	0939	3.900	0.1000	4.138	0.2184	0.0981	117	-1.09	019.53	0
019	Calcium (%)	0948	3.940	0.0400	4.138	0.2184	0.0981	117	-0.91	019.31	0
019	Calcium (%)	2129	3.948	0.1790	4.138	0.2184	0.0981	117	-0.87	019.42	0
019	Calcium (%)	0144	3.960	0.0800	4.138	0.2184	0.0981	117	-0.82	019.41	0
019	Calcium (%)	0164	3.965	0.1700	4.138	0.2184	0.0981	117	-0.79	019.44	0
019	Calcium (%)	0032	3.966	0.1000	4.138	0.2184	0.0981	117	-0.79	019.43	0
019	Calcium (%)	2109	3.979	0.2700	4.138	0.2184	0.0981	117	-0.73	019.00	0
019	Calcium (%)	0018	3.980	0.1600	4.138	0.2184	0.0981	117	-0.73	019.41	0
019	Calcium (%)	0918	3.983	0.0010	4.138	0.2184	0.0981	117	-0.71	019.43	0
019	Calcium (%)	0017	3.984	0.0611	4.138	0.2184	0.0981	117	-0.71	019.43	0
019	Calcium (%)	0278	3.990	0.0400	4.138	0.2184	0.0981	117	-0.68	019.42	0
019	Calcium (%)	0354	3.995	0.0100	4.138	0.2184	0.0981	117	-0.66	019.41	0
019	Calcium (%)	0870	3.996	0.0010	4.138	0.2184	0.0981	117	-0.65	019.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
019	Calcium (%)	2053	4.000	0.0000	4.138	0.2184	0.0981	117	-0.63	019.42	0
019	Calcium (%)	0010	4.005	0.0700	4.138	0.2184	0.0981	117	-0.61	019.43	0
019	Calcium (%)	0674	4.010	0.0000	4.138	0.2184	0.0981	117	-0.59	019.31	0
019	Calcium (%)	0357	4.012	0.0590	4.138	0.2184	0.0981	117	-0.58	019.42	0
019	Calcium (%)	0345	4.014	0.0340	4.138	0.2184	0.0981	117	-0.57	019.43	0
019	Calcium (%)	0878	4.020	0.1400	4.138	0.2184	0.0981	117	-0.54	019.41	0
019	Calcium (%)	0968	4.021	0.0450	4.138	0.2184	0.0981	117	-0.54	019.43	0
019	Calcium (%)	0139	4.045	0.0100	4.138	0.2184	0.0981	117	-0.43	019.41	0
019	Calcium (%)	0366	4.050	0.1600	4.138	0.2184	0.0981	117	-0.40	019.42	0
019	Calcium (%)	0083	4.050	0.0400	4.138	0.2184	0.0981	117	-0.40	019.43	0
019	Calcium (%)	2009	4.052	0.0179	4.138	0.2184	0.0981	117	-0.39	019.08	0
019	Calcium (%)	0675	4.055	0.0500	4.138	0.2184	0.0981	117	-0.38	019.43	0
019	Calcium (%)	0536	4.060	0.0600	4.138	0.2184	0.0981	117	-0.36	019.31	0
019	Calcium (%)	2181	4.065	0.0500	4.138	0.2184	0.0981	117	-0.34	019.00	0
019	Calcium (%)	0407	4.065	0.4004	4.138	0.2184	0.0981	117	-0.33	019.43	0
019	Calcium (%)	0242	4.070	0.0200	4.138	0.2184	0.0981	117	-0.31	019.99	0
019	Calcium (%)	0208	4.078	0.5030	4.138	0.2184	0.0981	117	-0.28	019.31	0
019	Calcium (%)	0169	4.080	0.0400	4.138	0.2184	0.0981	117	-0.27	019.32	0
019	Calcium (%)	0598	4.083	0.0440	4.138	0.2184	0.0981	117	-0.25	019.41	0
019	Calcium (%)	0878	4.085	0.0100	4.138	0.2184	0.0981	117	-0.24	019.00	0
019	Calcium (%)	0148	4.085	0.1100	4.138	0.2184	0.0981	117	-0.24	019.41	0
019	Calcium (%)	0142	4.087	0.1165	4.138	0.2184	0.0981	117	-0.23	019.31	0
019	Calcium (%)	0676	4.090	0.0000	4.138	0.2184	0.0981	117	-0.22	019.99	0
019	Calcium (%)	0692	4.095	0.0100	4.138	0.2184	0.0981	117	-0.20	019.42	0
019	Calcium (%)	0175	4.095	0.0500	4.138	0.2184	0.0981	117	-0.20	019.31	0
019	Calcium (%)	0297	4.095	0.2100	4.138	0.2184	0.0981	117	-0.20	019.43	0
019	Calcium (%)	0687	4.100	0.2000	4.138	0.2184	0.0981	117	-0.18	019.31	0
019	Calcium (%)	0848	4.100	0.1000	4.138	0.2184	0.0981	117	-0.18	019.41	0
019	Calcium (%)	0505	4.105	0.1900	4.138	0.2184	0.0981	117	-0.15	019.43	0
019	Calcium (%)	0689	4.125	0.1300	4.138	0.2184	0.0981	117	-0.06	019.08	0
019	Calcium (%)	0668	4.140	0.0000	4.138	0.2184	0.0981	117	0.01	019.43	0
019	Calcium (%)	0049	4.145	0.0900	4.138	0.2184	0.0981	117	0.03	019.41	0
019	Calcium (%)	0208	4.149	0.0530	4.138	0.2184	0.0981	117	0.05	019.41	0
019	Calcium (%)	0265	4.150	0.3400	4.138	0.2184	0.0981	117	0.05	019.42	0
019	Calcium (%)	0033	4.150	0.2400	4.138	0.2184	0.0981	117	0.05	019.43	0
019	Calcium (%)	2188	4.160	0.0000	4.138	0.2184	0.0981	117	0.10	019.08	0
019	Calcium (%)	0610	4.160	0.0200	4.138	0.2184	0.0981	117	0.10	019.43	0
019	Calcium (%)	0229	4.161	0.0900	4.138	0.2184	0.0981	117	0.10	019.41	0
019	Calcium (%)	0043	4.162	0.0540	4.138	0.2184	0.0981	117	0.11	019.43	0
019	Calcium (%)	0685	4.175	0.0100	4.138	0.2184	0.0981	117	0.17	019.41	0
019	Calcium (%)	0512	4.175	0.0900	4.138	0.2184	0.0981	117	0.17	019.43	0
019	Calcium (%)	0202	4.175	0.1100	4.138	0.2184	0.0981	117	0.17	019.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
019	Calcium (%)	0590	4.180	0.0200	4.138	0.2184	0.0981	117	0.19	019.08	0
019	Calcium (%)	0045	4.195	0.0700	4.138	0.2184	0.0981	117	0.26	019.42	0
019	Calcium (%)	0019	4.205	0.0500	4.138	0.2184	0.0981	117	0.31	019.41	0
019	Calcium (%)	0042	4.205	0.4300	4.138	0.2184	0.0981	117	0.31	019.42	0
019	Calcium (%)	0407	4.224	0.2549	4.138	0.2184	0.0981	117	0.39	019.41	0
019	Calcium (%)	0881	4.237	0.0360	4.138	0.2184	0.0981	117	0.45	019.00	0
019	Calcium (%)	2196	4.240	0.0000	4.138	0.2184	0.0981	117	0.47	019.08	0
019	Calcium (%)	0510	4.240	0.0200	4.138	0.2184	0.0981	117	0.47	019.43	0
019	Calcium (%)	0590	4.240	0.1400	4.138	0.2184	0.0981	117	0.47	019.99	0
019	Calcium (%)	0964	4.244	0.0423	4.138	0.2184	0.0981	117	0.48	019.43	0
019	Calcium (%)	0160	4.245	0.1900	4.138	0.2184	0.0981	117	0.49	019.44	0
019	Calcium (%)	2196	4.250	0.0000	4.138	0.2184	0.0981	117	0.51	019.31	0
019	Calcium (%)	0560	4.250	0.0549	4.138	0.2184	0.0981	117	0.51	019.52	0
019	Calcium (%)	2006	4.264	0.0060	4.138	0.2184	0.0981	117	0.58	019.00	0
019	Calcium (%)	0074	4.265	0.1100	4.138	0.2184	0.0981	117	0.58	019.41	0
019	Calcium (%)	0651	4.274	0.0596	4.138	0.2184	0.0981	117	0.62	019.00	0
019	Calcium (%)	0006	4.289	0.2020	4.138	0.2184	0.0981	117	0.69	019.43	0
019	Calcium (%)	0529	4.290	0.1200	4.138	0.2184	0.0981	117	0.69	019.31	0
019	Calcium (%)	0358	4.290	0.1200	4.138	0.2184	0.0981	117	0.69	019.41	0
019	Calcium (%)	0653	4.300	0.0120	4.138	0.2184	0.0981	117	0.74	019.41	0
019	Calcium (%)	0036	4.309	0.1164	4.138	0.2184	0.0981	117	0.78	019.03	0
019	Calcium (%)	0693	4.322	0.3040	4.138	0.2184	0.0981	117	0.84	019.42	0
019	Calcium (%)	2193	4.325	0.0100	4.138	0.2184	0.0981	117	0.85	019.08	0
019	Calcium (%)	0622	4.326	0.0440	4.138	0.2184	0.0981	117	0.86	019.31	0
019	Calcium (%)	2192	4.335	0.0700	4.138	0.2184	0.0981	117	0.90	019.43	0
019	Calcium (%)	0027	4.356	0.0620	4.138	0.2184	0.0981	117	1.00	019.43	0
019	Calcium (%)	0098	4.375	0.2300	4.138	0.2184	0.0981	117	1.08	019.41	0
019	Calcium (%)	0682	4.380	0.0000	4.138	0.2184	0.0981	117	1.11	019.43	0
019	Calcium (%)	0100	4.385	0.0300	4.138	0.2184	0.0981	117	1.13	019.99	0
019	Calcium (%)	0043	4.387	0.1000	4.138	0.2184	0.0981	117	1.14	019.00	0
019	Calcium (%)	2259	4.389	0.0150	4.138	0.2184	0.0981	117	1.15	019.41	0
019	Calcium (%)	0037	4.391	0.0500	4.138	0.2184	0.0981	117	1.16	019.42	0
019	Calcium (%)	0226	4.395	0.0100	4.138	0.2184	0.0981	117	1.18	019.41	0
019	Calcium (%)	0098	4.410	0.0200	4.138	0.2184	0.0981	117	1.24	019.43	0
019	Calcium (%)	0047	4.415	0.3100	4.138	0.2184	0.0981	117	1.27	019.52	0
019	Calcium (%)	0726	4.430	0.0600	4.138	0.2184	0.0981	117	1.34	019.42	0
019	Calcium (%)	0190	4.440	0.0800	4.138	0.2184	0.0981	117	1.38	019.42	0
019	Calcium (%)	2199	4.470	0.0200	4.138	0.2184	0.0981	117	1.52	019.08	0
019	Calcium (%)	0619	4.480	0.1600	4.138	0.2184	0.0981	117	1.56	019.41	0
019	Calcium (%)	0001	4.501	0.0510	4.138	0.2184	0.0981	117	1.66	019.31	0
019	Calcium (%)	0263	4.571	0.0580	4.138	0.2184	0.0981	117	1.98	019.42	0
019	Calcium (%)	2006	4.573	0.0680	4.138	0.2184	0.0981	117	1.99	019.09	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
019	Calcium (%)	0035	4.575	0.1500	4.138	0.2184	0.0981	117	2.00	019.42	0
019	Calcium (%)	0035	4.595	0.1100	4.138	0.2184	0.0981	117	2.09	019.43	0
019	Calcium (%)	0353	4.705	0.0100	4.138	0.2184	0.0981	117	2.59	019.43	0
019	Calcium (%)	0520	4.880	0.2200	4.138	0.2184	0.0981	117	3.40	019.41	0
019	Calcium (%)	0003	4.115	0.5700	4.138	0.2184	0.0981	117	-0.11	019.41	1
019	Calcium (%)	0921	41.52	0.5700	4.138	0.2184	0.0981	117	171.17	019.31	2
021	Cobalt (ppm)	0619	0.6460	0.0260	1.159	0.3053	0.0978	23	-1.68	021.41	0
021	Cobalt (ppm)	0939	0.6600	0.0600	1.159	0.3053	0.0978	23	-1.63	021.31	0
021	Cobalt (ppm)	0689	0.8000	0.2000	1.159	0.3053	0.0978	23	-1.18	021.31	0
021	Cobalt (ppm)	0693	0.8170	0.2120	1.159	0.3053	0.0978	23	-1.12	021.42	0
021	Cobalt (ppm)	0171	0.8900	0.0080	1.159	0.3053	0.0978	23	-0.88	021.41	0
021	Cobalt (ppm)	0910	0.9850	0.0500	1.159	0.3053	0.0978	23	-0.57	021.52	0
021	Cobalt (ppm)	0682	1.000	0.0000	1.159	0.3053	0.0978	23	-0.52	021.43	0
021	Cobalt (ppm)	0345	1.030	0.0600	1.159	0.3053	0.0978	23	-0.42	021.43	0
021	Cobalt (ppm)	0918	1.034	0.0041	1.159	0.3053	0.0978	23	-0.41	021.53	0
021	Cobalt (ppm)	0560	1.130	0.0400	1.159	0.3053	0.0978	23	-0.09	021.52	0
021	Cobalt (ppm)	0047	1.155	0.1100	1.159	0.3053	0.0978	23	-0.01	021.52	0
021	Cobalt (ppm)	0038	1.171	0.1490	1.159	0.3053	0.0978	23	0.04	021.43	0
021	Cobalt (ppm)	0186	1.190	0.1600	1.159	0.3053	0.0978	23	0.10	021.52	0
021	Cobalt (ppm)	0939	1.240	0.0000	1.159	0.3053	0.0978	23	0.27	021.53	0
021	Cobalt (ppm)	0169	1.255	0.0500	1.159	0.3053	0.0978	23	0.32	021.43	0
021	Cobalt (ppm)	0045	1.315	0.1700	1.159	0.3053	0.0978	23	0.51	021.42	0
021	Cobalt (ppm)	0407	1.338	0.0435	1.159	0.3053	0.0978	23	0.59	021.41	0
021	Cobalt (ppm)	0407	1.358	0.0518	1.159	0.3053	0.0978	23	0.65	021.43	0
021	Cobalt (ppm)	0964	1.373	0.3368	1.159	0.3053	0.0978	23	0.70	021.43	0
021	Cobalt (ppm)	0510	1.450	0.0000	1.159	0.3053	0.0978	23	0.95	021.43	0
021	Cobalt (ppm)	0148	1.485	0.0700	1.159	0.3053	0.0978	23	1.07	021.41	0
021	Cobalt (ppm)	0164	1.650	0.1000	1.159	0.3053	0.0978	23	1.61	021.31	0
021	Cobalt (ppm)	0353	1.879	0.3485	1.159	0.3053	0.0978	23	2.36	021.43	0
021	Cobalt (ppm)	2302	< 0.1		1.159	0.3053	0.0978	23		021.99	5
022	Copper (ppm)	2188	21.39	0.0000	27.54	2.328	1.357	87	-2.64	022.31	0
022	Copper (ppm)	0278	22.85	1.500	27.54	2.328	1.357	87	-2.01	022.42	0
022	Copper (ppm)	0848	23.46	0.1100	27.54	2.328	1.357	87	-1.75	022.41	0
022	Copper (ppm)	0345	23.75	1.500	27.54	2.328	1.357	87	-1.63	022.43	0
022	Copper (ppm)	0407	23.80	0.7383	27.54	2.328	1.357	87	-1.61	022.43	0
022	Copper (ppm)	0968	24.00	2.000	27.54	2.328	1.357	87	-1.52	022.43	0
022	Copper (ppm)	0939	24.32	0.0200	27.54	2.328	1.357	87	-1.38	022.53	0
022	Copper (ppm)	0171	24.46	1.640	27.54	2.328	1.357	87	-1.32	022.41	0
022	Copper (ppm)	0536	24.48	0.6500	27.54	2.328	1.357	87	-1.31	022.31	0
022	Copper (ppm)	0242	24.50	3.000	27.54	2.328	1.357	87	-1.30	022.99	0
022	Copper (ppm)	0653	24.56	1.270	27.54	2.328	1.357	87	-1.28	022.41	0
022	Copper (ppm)	0263	24.84	0.4933	27.54	2.328	1.357	87	-1.16	022.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
022	Copper (ppm)	0407	24.95	0.3716	27.54	2.328	1.357	87	-1.11	022.41	0
022	Copper (ppm)	0511	25.00	0.0000	27.54	2.328	1.357	87	-1.09	022.41	0
022	Copper (ppm)	0353	25.00	2.000	27.54	2.328	1.357	87	-1.09	022.43	0
022	Copper (ppm)	0169	25.05	0.3000	27.54	2.328	1.357	87	-1.07	022.43	0
022	Copper (ppm)	0098	25.30	0.6300	27.54	2.328	1.357	87	-0.96	022.43	0
022	Copper (ppm)	0186	25.75	0.3000	27.54	2.328	1.357	87	-0.77	022.42	0
022	Copper (ppm)	0939	25.79	0.0900	27.54	2.328	1.357	87	-0.75	022.31	0
022	Copper (ppm)	0563	25.88	1.113	27.54	2.328	1.357	87	-0.71	022.31	0
022	Copper (ppm)	0032	26.00	1.190	27.54	2.328	1.357	87	-0.66	022.43	0
022	Copper (ppm)	0870	26.00	1.110	27.54	2.328	1.357	87	-0.66	022.42	0
022	Copper (ppm)	0878	26.00	2.000	27.54	2.328	1.357	87	-0.66	022.41	0
022	Copper (ppm)	0910	26.00	2.000	27.54	2.328	1.357	87	-0.66	022.41	0
022	Copper (ppm)	0357	26.30	0.5470	27.54	2.328	1.357	87	-0.53	022.42	0
022	Copper (ppm)	2196	26.30	0.0000	27.54	2.328	1.357	87	-0.53	022.31	0
022	Copper (ppm)	0098	26.45	5.010	27.54	2.328	1.357	87	-0.47	022.41	0
022	Copper (ppm)	0619	26.55	0.5000	27.54	2.328	1.357	87	-0.42	022.41	0
022	Copper (ppm)	0208	26.64	0.1900	27.54	2.328	1.357	87	-0.39	022.41	0
022	Copper (ppm)	0202	26.79	1.910	27.54	2.328	1.357	87	-0.32	022.43	0
022	Copper (ppm)	0010	26.85	0.5000	27.54	2.328	1.357	87	-0.29	022.33	0
022	Copper (ppm)	2192	26.92	3.120	27.54	2.328	1.357	87	-0.26	022.43	0
022	Copper (ppm)	0035	27.00	0.0000	27.54	2.328	1.357	87	-0.23	022.42	0
022	Copper (ppm)	0083	27.00	0.0000	27.54	2.328	1.357	87	-0.23	022.43	0
022	Copper (ppm)	0226	27.15	0.6500	27.54	2.328	1.357	87	-0.17	022.41	0
022	Copper (ppm)	0186	27.20	0.2000	27.54	2.328	1.357	87	-0.14	022.52	0
022	Copper (ppm)	0354	27.25	0.0400	27.54	2.328	1.357	87	-0.12	022.41	0
022	Copper (ppm)	0512	27.25	2.780	27.54	2.328	1.357	87	-0.12	022.43	0
022	Copper (ppm)	0004	27.31	0.9800	27.54	2.328	1.357	87	-0.10	022.41	0
022	Copper (ppm)	0668	27.45	2.300	27.54	2.328	1.357	87	-0.04	022.43	0
022	Copper (ppm)	0003	27.50	3.000	27.54	2.328	1.357	87	-0.02	022.41	0
022	Copper (ppm)	0074	27.50	1.000	27.54	2.328	1.357	87	-0.02	022.41	0
022	Copper (ppm)	0019	27.61	1.000	27.54	2.328	1.357	87	0.03	022.41	0
022	Copper (ppm)	0560	27.65	0.7000	27.54	2.328	1.357	87	0.05	022.52	0
022	Copper (ppm)	0049	27.79	4.560	27.54	2.328	1.357	87	0.11	022.41	0
022	Copper (ppm)	0208	28.00	2.400	27.54	2.328	1.357	87	0.20	022.31	0
022	Copper (ppm)	0689	28.00	0.2000	27.54	2.328	1.357	87	0.20	022.31	0
022	Copper (ppm)	0520	28.00	2.000	27.54	2.328	1.357	87	0.20	022.41	0
022	Copper (ppm)	0366	28.00	2.000	27.54	2.328	1.357	87	0.20	022.42	0
022	Copper (ppm)	0685	28.00	0.0400	27.54	2.328	1.357	87	0.20	022.42	0
022	Copper (ppm)	0035	28.00	2.000	27.54	2.328	1.357	87	0.20	022.43	0
022	Copper (ppm)	0100	28.00	4.000	27.54	2.328	1.357	87	0.20	022.99	0
022	Copper (ppm)	0622	28.05	1.141	27.54	2.328	1.357	87	0.22	022.31	0
022	Copper (ppm)	0033	28.05	3.700	27.54	2.328	1.357	87	0.22	022.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
022	Copper (ppm)	0651	28.08	1.300	27.54	2.328	1.357	87	0.23	022.31	0
022	Copper (ppm)	2129	28.18	1.300	27.54	2.328	1.357	87	0.28	022.42	0
022	Copper (ppm)	0726	28.22	0.5400	27.54	2.328	1.357	87	0.29	022.42	0
022	Copper (ppm)	0038	28.28	3.827	27.54	2.328	1.357	87	0.32	022.43	0
022	Copper (ppm)	0675	28.29	0.7500	27.54	2.328	1.357	87	0.32	022.43	0
022	Copper (ppm)	0229	28.39	0.8000	27.54	2.328	1.357	87	0.37	022.41	0
022	Copper (ppm)	0682	28.50	0.0000	27.54	2.328	1.357	87	0.41	022.43	0
022	Copper (ppm)	0160	28.55	3.300	27.54	2.328	1.357	87	0.44	022.44	0
022	Copper (ppm)	0405	28.62	0.9000	27.54	2.328	1.357	87	0.47	022.41	0
022	Copper (ppm)	0693	28.83	0.3110	27.54	2.328	1.357	87	0.56	022.42	0
022	Copper (ppm)	0692	28.85	5.700	27.54	2.328	1.357	87	0.56	022.42	0
022	Copper (ppm)	0045	28.90	2.600	27.54	2.328	1.357	87	0.59	022.42	0
022	Copper (ppm)	0590	29.25	0.1000	27.54	2.328	1.357	87	0.74	022.99	0
022	Copper (ppm)	0610	29.35	1.100	27.54	2.328	1.357	87	0.78	022.43	0
022	Copper (ppm)	0510	29.50	1.000	27.54	2.328	1.357	87	0.84	022.43	0
022	Copper (ppm)	0164	29.50	1.000	27.54	2.328	1.357	87	0.84	022.44	0
022	Copper (ppm)	0918	29.59	1.280	27.54	2.328	1.357	87	0.88	022.43	0
022	Copper (ppm)	0529	29.60	4.400	27.54	2.328	1.357	87	0.89	022.31	0
022	Copper (ppm)	0017	29.82	0.6875	27.54	2.328	1.357	87	0.98	022.43	0
022	Copper (ppm)	0948	29.87	0.6940	27.54	2.328	1.357	87	1.00	022.33	0
022	Copper (ppm)	0265	29.90	5.400	27.54	2.328	1.357	87	1.02	022.42	0
022	Copper (ppm)	0175	30.00	0.0000	27.54	2.328	1.357	87	1.06	022.31	0
022	Copper (ppm)	0190	30.28	0.0300	27.54	2.328	1.357	87	1.18	022.42	0
022	Copper (ppm)	0148	30.40	0.2000	27.54	2.328	1.357	87	1.23	022.41	0
022	Copper (ppm)	0297	30.50	1.000	27.54	2.328	1.357	87	1.27	022.43	0
022	Copper (ppm)	0598	30.51	0.0900	27.54	2.328	1.357	87	1.28	022.41	0
022	Copper (ppm)	0964	31.41	2.599	27.54	2.328	1.357	87	1.67	022.43	0
022	Copper (ppm)	0358	31.54	1.630	27.54	2.328	1.357	87	1.72	022.41	0
022	Copper (ppm)	2053	31.75	0.7000	27.54	2.328	1.357	87	1.81	022.42	0
022	Copper (ppm)	0027	32.20	1.493	27.54	2.328	1.357	87	2.01	022.43	0
022	Copper (ppm)	0505	33.09	1.040	27.54	2.328	1.357	87	2.39	022.43	0
022	Copper (ppm)	0047	33.64	0.2800	27.54	2.328	1.357	87	2.62	022.52	0
022	Copper (ppm)	0042	34.25	1.500	27.54	2.328	1.357	87	2.88	022.42	0
022	Copper (ppm)	2062	33.10	7.600	27.54	2.328	1.357	87	2.39	022.31	1
022	Copper (ppm)	0037	38.45	10.70	27.54	2.328	1.357	87	4.69	022.42	1
024	Iodine (ppm)	0939	5.360	0.4600				1		024.53	0
025	Iron (ppm)	2053	134.0	4.000	273.9	29.46	9.029	80	-4.75	025.42	0
025	Iron (ppm)	0405	205.4	9.290	273.9	29.46	9.029	80	-2.32	025.41	0
025	Iron (ppm)	0265	212.0	22.00	273.9	29.46	9.029	80	-2.10	025.42	0
025	Iron (ppm)	0921	214.7	21.98	273.9	29.46	9.029	80	-2.01	025.31	0
025	Iron (ppm)	0047	215.0	3.710	273.9	29.46	9.029	80	-2.00	025.52	0
025	Iron (ppm)	0037	217.1	13.30	273.9	29.46	9.029	80	-1.93	025.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
025	Iron (ppm)	0278	237.0	12.00	273.9	29.46	9.029	80	-1.25	025.42	0
025	Iron (ppm)	0354	239.2	0.3800	273.9	29.46	9.029	80	-1.18	025.41	0
025	Iron (ppm)	0242	242.0	10.00	273.9	29.46	9.029	80	-1.08	025.99	0
025	Iron (ppm)	0407	242.4	12.63	273.9	29.46	9.029	80	-1.07	025.41	0
025	Iron (ppm)	0692	243.0	6.000	273.9	29.46	9.029	80	-1.05	025.42	0
025	Iron (ppm)	0345	243.0	6.000	273.9	29.46	9.029	80	-1.05	025.43	0
025	Iron (ppm)	0675	244.9	4.010	273.9	29.46	9.029	80	-0.99	025.43	0
025	Iron (ppm)	0653	246.9	4.800	273.9	29.46	9.029	80	-0.92	025.41	0
025	Iron (ppm)	0123	247.0	2.000	273.9	29.46	9.029	80	-0.91	025.41	0
025	Iron (ppm)	0353	247.0	18.00	273.9	29.46	9.029	80	-0.91	025.43	0
025	Iron (ppm)	0160	248.0	14.00	273.9	29.46	9.029	80	-0.88	025.99	0
025	Iron (ppm)	0100	251.0	2.000	273.9	29.46	9.029	80	-0.78	025.99	0
025	Iron (ppm)	0726	252.0	0.5000	273.9	29.46	9.029	80	-0.75	025.42	0
025	Iron (ppm)	0366	253.0	12.00	273.9	29.46	9.029	80	-0.71	025.42	0
025	Iron (ppm)	0407	253.3	7.848	273.9	29.46	9.029	80	-0.70	025.43	0
025	Iron (ppm)	0878	253.5	7.000	273.9	29.46	9.029	80	-0.69	025.41	0
025	Iron (ppm)	0169	254.0	2.000	273.9	29.46	9.029	80	-0.68	025.43	0
025	Iron (ppm)	0042	254.5	13.00	273.9	29.46	9.029	80	-0.66	025.42	0
025	Iron (ppm)	0870	256.3	2.800	273.9	29.46	9.029	80	-0.60	025.42	0
025	Iron (ppm)	0598	260.1	1.300	273.9	29.46	9.029	80	-0.47	025.41	0
025	Iron (ppm)	0263	261.1	3.870	273.9	29.46	9.029	80	-0.43	025.42	0
025	Iron (ppm)	0910	262.5	27.00	273.9	29.46	9.029	80	-0.39	025.41	0
025	Iron (ppm)	0045	262.5	1.000	273.9	29.46	9.029	80	-0.39	025.42	0
025	Iron (ppm)	0610	264.5	7.000	273.9	29.46	9.029	80	-0.32	025.43	0
025	Iron (ppm)	0511	265.0	4.000	273.9	29.46	9.029	80	-0.30	025.41	0
025	Iron (ppm)	2129	266.3	1.600	273.9	29.46	9.029	80	-0.26	025.42	0
025	Iron (ppm)	0505	267.2	3.290	273.9	29.46	9.029	80	-0.23	025.43	0
025	Iron (ppm)	0689	267.5	3.000	273.9	29.46	9.029	80	-0.22	025.31	0
025	Iron (ppm)	2188	268.0	0.0000	273.9	29.46	9.029	80	-0.20	025.31	0
025	Iron (ppm)	0297	270.5	13.00	273.9	29.46	9.029	80	-0.12	025.43	0
025	Iron (ppm)	0098	272.0	15.20	273.9	29.46	9.029	80	-0.07	025.41	0
025	Iron (ppm)	0190	273.6	3.020	273.9	29.46	9.029	80	-0.01	025.42	0
025	Iron (ppm)	0003	276.0	28.00	273.9	29.46	9.029	80	0.07	025.41	0
025	Iron (ppm)	0693	277.0	10.27	273.9	29.46	9.029	80	0.10	025.42	0
025	Iron (ppm)	0939	277.0	2.000	273.9	29.46	9.029	80	0.10	025.31	0
025	Iron (ppm)	0083	277.0	22.00	273.9	29.46	9.029	80	0.10	025.43	0
025	Iron (ppm)	0918	277.1	2.570	273.9	29.46	9.029	80	0.11	025.43	0
025	Iron (ppm)	0668	277.5	9.000	273.9	29.46	9.029	80	0.12	025.43	0
025	Iron (ppm)	0619	278.0	4.000	273.9	29.46	9.029	80	0.14	025.41	0
025	Iron (ppm)	0939	278.0	2.000	273.9	29.46	9.029	80	0.14	025.53	0
025	Iron (ppm)	0004	278.4	32.64	273.9	29.46	9.029	80	0.15	025.41	0
025	Iron (ppm)	0164	278.5	3.000	273.9	29.46	9.029	80	0.16	025.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
025	Iron (ppm)	0622	282.4	4.814	273.9	29.46	9.029	80	0.29	025.31	0
025	Iron (ppm)	0651	282.4	4.814	273.9	29.46	9.029	80	0.29	025.31	0
025	Iron (ppm)	0968	282.5	1.000	273.9	29.46	9.029	80	0.29	025.43	0
025	Iron (ppm)	0098	283.6	0.9000	273.9	29.46	9.029	80	0.33	025.43	0
025	Iron (ppm)	0226	283.7	12.22	273.9	29.46	9.029	80	0.33	025.41	0
025	Iron (ppm)	0032	285.0	13.60	273.9	29.46	9.029	80	0.38	025.43	0
025	Iron (ppm)	0563	286.0	4.423	273.9	29.46	9.029	80	0.41	025.31	0
025	Iron (ppm)	0171	286.2	7.200	273.9	29.46	9.029	80	0.42	025.41	0
025	Iron (ppm)	2196	287.9	0.0000	273.9	29.46	9.029	80	0.47	025.31	0
025	Iron (ppm)	0208	288.6	23.60	273.9	29.46	9.029	80	0.50	025.41	0
025	Iron (ppm)	0074	290.5	7.000	273.9	29.46	9.029	80	0.56	025.41	0
025	Iron (ppm)	0033	290.5	7.000	273.9	29.46	9.029	80	0.56	025.43	0
025	Iron (ppm)	0019	294.2	5.700	273.9	29.46	9.029	80	0.69	025.41	0
025	Iron (ppm)	0229	294.7	5.400	273.9	29.46	9.029	80	0.71	025.41	0
025	Iron (ppm)	0017	295.1	25.03	273.9	29.46	9.029	80	0.72	025.43	0
025	Iron (ppm)	0529	295.3	6.500	273.9	29.46	9.029	80	0.72	025.31	0
025	Iron (ppm)	0682	297.9	0.0000	273.9	29.46	9.029	80	0.81	025.43	0
025	Iron (ppm)	0208	300.5	17.00	273.9	29.46	9.029	80	0.90	025.31	0
025	Iron (ppm)	0510	300.5	1.000	273.9	29.46	9.029	80	0.90	025.43	0
025	Iron (ppm)	0358	302.2	6.890	273.9	29.46	9.029	80	0.96	025.41	0
025	Iron (ppm)	2192	307.8	12.98	273.9	29.46	9.029	80	1.15	025.43	0
025	Iron (ppm)	0964	309.2	17.10	273.9	29.46	9.029	80	1.20	025.43	0
025	Iron (ppm)	0148	312.0	12.00	273.9	29.46	9.029	80	1.29	025.41	0
025	Iron (ppm)	0038	312.5	15.00	273.9	29.46	9.029	80	1.31	025.43	0
025	Iron (ppm)	0049	314.3	32.11	273.9	29.46	9.029	80	1.37	025.41	0
025	Iron (ppm)	0520	320.5	5.000	273.9	29.46	9.029	80	1.58	025.41	0
025	Iron (ppm)	0512	325.8	13.40	273.9	29.46	9.029	80	1.76	025.43	0
025	Iron (ppm)	0175	329.0	2.000	273.9	29.46	9.029	80	1.87	025.31	0
025	Iron (ppm)	2062	336.5	25.90	273.9	29.46	9.029	80	2.12	025.31	0
025	Iron (ppm)	0035	340.0	12.00	273.9	29.46	9.029	80	2.24	025.43	0
025	Iron (ppm)	0035	351.0	6.000	273.9	29.46	9.029	80	2.62	025.42	0
025	Iron (ppm)	0536	358.5	4.700	273.9	29.46	9.029	80	2.87	025.31	0
025	Iron (ppm)	0560	242.9	56.40	273.9	29.46	9.029	80	-1.05	025.52	1
027	Magnesium (%)	0407	0.1946	0.0046	0.2245	0.0128	0.0054	87	-2.33	027.43	0
027	Magnesium (%)	0407	0.1966	0.0045	0.2245	0.0128	0.0054	87	-2.17	027.41	0
027	Magnesium (%)	0202	0.1980	0.0040	0.2245	0.0128	0.0054	87	-2.06	027.43	0
027	Magnesium (%)	0083	0.2000	0.0000	0.2245	0.0128	0.0054	87	-1.90	027.43	0
027	Magnesium (%)	0511	0.2050	0.0100	0.2245	0.0128	0.0054	87	-1.52	027.41	0
027	Magnesium (%)	0505	0.2050	0.0100	0.2245	0.0128	0.0054	87	-1.52	027.43	0
027	Magnesium (%)	0529	0.2060	0.0020	0.2245	0.0128	0.0054	87	-1.44	027.31	0
027	Magnesium (%)	0870	0.2062	0.0068	0.2245	0.0128	0.0054	87	-1.42	027.42	0
027	Magnesium (%)	0265	0.2100	0.0000	0.2245	0.0128	0.0054	87	-1.13	027.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
027	Magnesium (%)	0510	0.2100	0.0000	0.2245	0.0128	0.0054	87	-1.13	027.43	0
027	Magnesium (%)	0590	0.2100	0.0000	0.2245	0.0128	0.0054	87	-1.13	027.99	0
027	Magnesium (%)	0047	0.2100	0.0200	0.2245	0.0128	0.0054	87	-1.13	027.52	0
027	Magnesium (%)	0032	0.2105	0.0017	0.2245	0.0128	0.0054	87	-1.09	027.43	0
027	Magnesium (%)	0357	0.2106	0.0003	0.2245	0.0128	0.0054	87	-1.08	027.42	0
027	Magnesium (%)	2129	0.2120	0.0066	0.2245	0.0128	0.0054	87	-0.97	027.42	0
027	Magnesium (%)	0345	0.2125	0.0050	0.2245	0.0128	0.0054	87	-0.93	027.43	0
027	Magnesium (%)	0668	0.2130	0.0000	0.2245	0.0128	0.0054	87	-0.89	027.43	0
027	Magnesium (%)	0910	0.2150	0.0100	0.2245	0.0128	0.0054	87	-0.74	027.41	0
027	Magnesium (%)	0278	0.2150	0.0100	0.2245	0.0128	0.0054	87	-0.74	027.42	0
027	Magnesium (%)	0366	0.2150	0.0100	0.2245	0.0128	0.0054	87	-0.74	027.42	0
027	Magnesium (%)	0353	0.2150	0.0100	0.2245	0.0128	0.0054	87	-0.74	027.43	0
027	Magnesium (%)	0229	0.2155	0.0050	0.2245	0.0128	0.0054	87	-0.70	027.41	0
027	Magnesium (%)	0560	0.2172	0.0142	0.2245	0.0128	0.0054	87	-0.56	027.52	0
027	Magnesium (%)	2053	0.2175	0.0010	0.2245	0.0128	0.0054	87	-0.54	027.42	0
027	Magnesium (%)	0186	0.2190	0.0020	0.2245	0.0128	0.0054	87	-0.42	027.42	0
027	Magnesium (%)	0045	0.2195	0.0050	0.2245	0.0128	0.0054	87	-0.39	027.42	0
027	Magnesium (%)	0175	0.2200	0.0000	0.2245	0.0128	0.0054	87	-0.35	027.31	0
027	Magnesium (%)	0939	0.2200	0.0000	0.2245	0.0128	0.0054	87	-0.35	027.31	0
027	Magnesium (%)	0003	0.2200	0.0200	0.2245	0.0128	0.0054	87	-0.35	027.41	0
027	Magnesium (%)	0004	0.2200	0.0200	0.2245	0.0128	0.0054	87	-0.35	027.41	0
027	Magnesium (%)	0049	0.2200	0.0000	0.2245	0.0128	0.0054	87	-0.35	027.41	0
027	Magnesium (%)	0148	0.2200	0.0000	0.2245	0.0128	0.0054	87	-0.35	027.41	0
027	Magnesium (%)	0171	0.2200	0.0000	0.2245	0.0128	0.0054	87	-0.35	027.41	0
027	Magnesium (%)	0692	0.2200	0.0000	0.2245	0.0128	0.0054	87	-0.35	027.42	0
027	Magnesium (%)	0035	0.2200	0.0000	0.2245	0.0128	0.0054	87	-0.35	027.43	0
027	Magnesium (%)	0610	0.2200	0.0000	0.2245	0.0128	0.0054	87	-0.35	027.43	0
027	Magnesium (%)	0675	0.2200	0.0000	0.2245	0.0128	0.0054	87	-0.35	027.43	0
027	Magnesium (%)	0242	0.2200	0.0000	0.2245	0.0128	0.0054	87	-0.35	027.99	0
027	Magnesium (%)	0968	0.2225	0.0030	0.2245	0.0128	0.0054	87	-0.15	027.43	0
027	Magnesium (%)	0037	0.2235	0.0010	0.2245	0.0128	0.0054	87	-0.07	027.42	0
027	Magnesium (%)	0598	0.2236	0.0047	0.2245	0.0128	0.0054	87	-0.07	027.41	0
027	Magnesium (%)	0142	0.2239	0.0100	0.2245	0.0128	0.0054	87	-0.04	027.31	0
027	Magnesium (%)	0208	0.2245	0.0030	0.2245	0.0128	0.0054	87	0.00	027.41	0
027	Magnesium (%)	0186	0.2245	0.0010	0.2245	0.0128	0.0054	87	0.00	027.52	0
027	Magnesium (%)	0878	0.2250	0.0100	0.2245	0.0128	0.0054	87	0.04	027.41	0
027	Magnesium (%)	0685	0.2250	0.0100	0.2245	0.0128	0.0054	87	0.04	027.42	0
027	Magnesium (%)	0033	0.2250	0.0100	0.2245	0.0128	0.0054	87	0.04	027.43	0
027	Magnesium (%)	0098	0.2250	0.0100	0.2245	0.0128	0.0054	87	0.04	027.43	0
027	Magnesium (%)	0297	0.2250	0.0100	0.2245	0.0128	0.0054	87	0.04	027.43	0
027	Magnesium (%)	0160	0.2250	0.0100	0.2245	0.0128	0.0054	87	0.04	027.44	0
027	Magnesium (%)	0939	0.2250	0.0100	0.2245	0.0128	0.0054	87	0.04	027.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
027	Magnesium (%)	0622	0.2264	0.0009	0.2245	0.0128	0.0054	87	0.15	027.31	0
027	Magnesium (%)	0208	0.2265	0.0050	0.2245	0.0128	0.0054	87	0.16	027.31	0
027	Magnesium (%)	0027	0.2265	0.0010	0.2245	0.0128	0.0054	87	0.16	027.43	0
027	Magnesium (%)	0139	0.2272	0.0010	0.2245	0.0128	0.0054	87	0.21	027.41	0
027	Magnesium (%)	0017	0.2278	0.0023	0.2245	0.0128	0.0054	87	0.26	027.43	0
027	Magnesium (%)	0164	0.2285	0.0010	0.2245	0.0128	0.0054	87	0.32	027.44	0
027	Magnesium (%)	2196	0.2300	0.0000	0.2245	0.0128	0.0054	87	0.43	027.31	0
027	Magnesium (%)	0019	0.2300	0.0000	0.2245	0.0128	0.0054	87	0.43	027.41	0
027	Magnesium (%)	0035	0.2300	0.0000	0.2245	0.0128	0.0054	87	0.43	027.42	0
027	Magnesium (%)	0726	0.2300	0.0000	0.2245	0.0128	0.0054	87	0.43	027.42	0
027	Magnesium (%)	0074	0.2320	0.0020	0.2245	0.0128	0.0054	87	0.59	027.41	0
027	Magnesium (%)	0001	0.2324	0.0038	0.2245	0.0128	0.0054	87	0.62	027.31	0
027	Magnesium (%)	0619	0.2325	0.0030	0.2245	0.0128	0.0054	87	0.63	027.41	0
027	Magnesium (%)	0006	0.2340	0.0100	0.2245	0.0128	0.0054	87	0.74	027.43	0
027	Magnesium (%)	0536	0.2350	0.0100	0.2245	0.0128	0.0054	87	0.82	027.31	0
027	Magnesium (%)	0689	0.2350	0.0100	0.2245	0.0128	0.0054	87	0.82	027.31	0
027	Magnesium (%)	0169	0.2350	0.0100	0.2245	0.0128	0.0054	87	0.82	027.32	0
027	Magnesium (%)	0098	0.2350	0.0100	0.2245	0.0128	0.0054	87	0.82	027.41	0
027	Magnesium (%)	0354	0.2350	0.0100	0.2245	0.0128	0.0054	87	0.82	027.41	0
027	Magnesium (%)	0520	0.2350	0.0100	0.2245	0.0128	0.0054	87	0.82	027.41	0
027	Magnesium (%)	0693	0.2355	0.0030	0.2245	0.0128	0.0054	87	0.86	027.42	0
027	Magnesium (%)	0918	0.2365	0.0010	0.2245	0.0128	0.0054	87	0.94	027.43	0
027	Magnesium (%)	0512	0.2367	0.0006	0.2245	0.0128	0.0054	87	0.95	027.43	0
027	Magnesium (%)	0948	0.2370	0.0000	0.2245	0.0128	0.0054	87	0.98	027.33	0
027	Magnesium (%)	0042	0.2385	0.0030	0.2245	0.0128	0.0054	87	1.09	027.42	0
027	Magnesium (%)	0226	0.2400	0.0200	0.2245	0.0128	0.0054	87	1.21	027.41	0
027	Magnesium (%)	0563	0.2406	0.0030	0.2245	0.0128	0.0054	87	1.26	027.31	0
027	Magnesium (%)	0653	0.2410	0.0060	0.2245	0.0128	0.0054	87	1.29	027.41	0
027	Magnesium (%)	0405	0.2445	0.0010	0.2245	0.0128	0.0054	87	1.56	027.41	0
027	Magnesium (%)	2192	0.2450	0.0100	0.2245	0.0128	0.0054	87	1.60	027.43	0
027	Magnesium (%)	0100	0.2450	0.0100	0.2245	0.0128	0.0054	87	1.60	027.99	0
027	Magnesium (%)	0263	0.2484	0.0016	0.2245	0.0128	0.0054	87	1.87	027.42	0
027	Magnesium (%)	0358	0.2500	0.0200	0.2245	0.0128	0.0054	87	1.99	027.41	0
027	Magnesium (%)	0964	0.2547	0.0131	0.2245	0.0128	0.0054	87	2.35	027.43	0
027	Magnesium (%)	0682	0.2600	0.0000	0.2245	0.0128	0.0054	87	2.77	027.43	0
027	Magnesium (%)	2146	0.3750	0.0100	0.2245	0.0128	0.0054	87	11.73	027.31	0
027	Magnesium (%)	0190	0.2150	0.0300	0.2245	0.0128	0.0054	87	-0.74	027.42	1
027	Magnesium (%)	0038	0.2260	0.0340	0.2245	0.0128	0.0054	87	0.12	027.43	1
027	Magnesium (%)	2146	0.3950	0.0100	0.2245	0.0128	0.0054	87	13.29	027.51	2
028	Manganese (ppm)	0226	113.5	3.810	165.5	14.20	6.648	86	-3.66	028.41	0
028	Manganese (ppm)	0689	118.5	0.4000	165.5	14.20	6.648	86	-3.31	028.31	0
028	Manganese (ppm)	0354	121.1	0.7400	165.5	14.20	6.648	86	-3.12	028.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
028	Manganese (ppm)	0619	124.0	2.000	165.5	14.20	6.648	86	-2.92	028.41	0
028	Manganese (ppm)	0653	131.9	3.800	165.5	14.20	6.648	86	-2.36	028.41	0
028	Manganese (ppm)	0948	137.2	9.337	165.5	14.20	6.648	86	-1.99	028.33	0
028	Manganese (ppm)	0560	141.8	11.50	165.5	14.20	6.648	86	-1.67	028.52	0
028	Manganese (ppm)	0598	142.7	1.500	165.5	14.20	6.648	86	-1.61	028.41	0
028	Manganese (ppm)	0685	143.2	0.0100	165.5	14.20	6.648	86	-1.57	028.42	0
028	Manganese (ppm)	0878	144.5	3.000	165.5	14.20	6.648	86	-1.48	028.41	0
028	Manganese (ppm)	0004	146.1	25.09	165.5	14.20	6.648	86	-1.36	028.41	0
028	Manganese (ppm)	0405	148.1	1.500	165.5	14.20	6.648	86	-1.22	028.41	0
028	Manganese (ppm)	2188	149.0	0.0000	165.5	14.20	6.648	86	-1.16	028.31	0
028	Manganese (ppm)	0692	151.0	10.00	165.5	14.20	6.648	86	-1.02	028.42	0
028	Manganese (ppm)	2129	151.2	7.500	165.5	14.20	6.648	86	-1.01	028.42	0
028	Manganese (ppm)	0353	152.5	1.000	165.5	14.20	6.648	86	-0.91	028.43	0
028	Manganese (ppm)	0918	152.7	1.610	165.5	14.20	6.648	86	-0.90	028.43	0
028	Manganese (ppm)	0407	152.7	4.981	165.5	14.20	6.648	86	-0.90	028.43	0
028	Manganese (ppm)	0511	153.0	2.000	165.5	14.20	6.648	86	-0.88	028.41	0
028	Manganese (ppm)	0366	155.5	17.00	165.5	14.20	6.648	86	-0.70	028.42	0
028	Manganese (ppm)	0910	157.0	8.000	165.5	14.20	6.648	86	-0.60	028.41	0
028	Manganese (ppm)	0263	157.9	0.0400	165.5	14.20	6.648	86	-0.53	028.42	0
028	Manganese (ppm)	0939	158.0	2.000	165.5	14.20	6.648	86	-0.53	028.53	0
028	Manganese (ppm)	0407	158.2	1.749	165.5	14.20	6.648	86	-0.51	028.41	0
028	Manganese (ppm)	0505	160.0	27.50	165.5	14.20	6.648	86	-0.39	028.43	0
028	Manganese (ppm)	2053	160.0	2.000	165.5	14.20	6.648	86	-0.39	028.42	0
028	Manganese (ppm)	2196	160.3	0.0000	165.5	14.20	6.648	86	-0.36	028.31	0
028	Manganese (ppm)	0939	160.5	3.000	165.5	14.20	6.648	86	-0.35	028.31	0
028	Manganese (ppm)	0345	160.5	3.000	165.5	14.20	6.648	86	-0.35	028.43	0
028	Manganese (ppm)	0049	161.5	19.04	165.5	14.20	6.648	86	-0.28	028.41	0
028	Manganese (ppm)	0512	162.3	26.80	165.5	14.20	6.648	86	-0.22	028.43	0
028	Manganese (ppm)	0017	162.7	17.47	165.5	14.20	6.648	86	-0.20	028.43	0
028	Manganese (ppm)	0045	163.0	10.00	165.5	14.20	6.648	86	-0.17	028.42	0
028	Manganese (ppm)	0098	163.8	2.800	165.5	14.20	6.648	86	-0.12	028.41	0
028	Manganese (ppm)	0675	164.0	1.640	165.5	14.20	6.648	86	-0.10	028.43	0
028	Manganese (ppm)	0171	164.2	0.5000	165.5	14.20	6.648	86	-0.09	028.41	0
028	Manganese (ppm)	0001	164.5	8.000	165.5	14.20	6.648	86	-0.07	028.31	0
028	Manganese (ppm)	0242	164.5	5.000	165.5	14.20	6.648	86	-0.07	028.99	0
028	Manganese (ppm)	0208	165.1	6.100	165.5	14.20	6.648	86	-0.03	028.41	0
028	Manganese (ppm)	0098	166.0	12.70	165.5	14.20	6.648	86	0.03	028.43	0
028	Manganese (ppm)	0693	166.1	6.084	165.5	14.20	6.648	86	0.05	028.42	0
028	Manganese (ppm)	0870	166.7	11.40	165.5	14.20	6.648	86	0.09	028.42	0
028	Manganese (ppm)	0186	167.0	2.000	165.5	14.20	6.648	86	0.11	028.42	0
028	Manganese (ppm)	0033	167.0	4.000	165.5	14.20	6.648	86	0.11	028.43	0
028	Manganese (ppm)	0164	167.0	8.000	165.5	14.20	6.648	86	0.11	028.44	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
028	Manganese (ppm)	0229	167.2	4.900	165.5	14.20	6.648	86	0.12	028.41	0
028	Manganese (ppm)	0610	167.5	19.00	165.5	14.20	6.648	86	0.14	028.43	0
028	Manganese (ppm)	0668	167.5	7.000	165.5	14.20	6.648	86	0.14	028.43	0
028	Manganese (ppm)	0563	168.4	0.8943	165.5	14.20	6.648	86	0.21	028.31	0
028	Manganese (ppm)	0169	168.5	1.000	165.5	14.20	6.648	86	0.21	028.43	0
028	Manganese (ppm)	0358	168.7	10.78	165.5	14.20	6.648	86	0.23	028.41	0
028	Manganese (ppm)	0297	169.0	12.00	165.5	14.20	6.648	86	0.25	028.43	0
028	Manganese (ppm)	0074	169.5	15.00	165.5	14.20	6.648	86	0.28	028.41	0
028	Manganese (ppm)	0278	169.5	17.00	165.5	14.20	6.648	86	0.28	028.42	0
028	Manganese (ppm)	0083	169.5	1.000	165.5	14.20	6.648	86	0.28	028.43	0
028	Manganese (ppm)	0529	169.6	8.100	165.5	14.20	6.648	86	0.29	028.31	0
028	Manganese (ppm)	0019	169.8	4.500	165.5	14.20	6.648	86	0.30	028.41	0
028	Manganese (ppm)	0208	170.0	4.000	165.5	14.20	6.648	86	0.32	028.31	0
028	Manganese (ppm)	0160	170.0	2.000	165.5	14.20	6.648	86	0.32	028.44	0
028	Manganese (ppm)	0042	170.5	5.000	165.5	14.20	6.648	86	0.35	028.42	0
028	Manganese (ppm)	0186	171.0	2.000	165.5	14.20	6.648	86	0.39	028.52	0
028	Manganese (ppm)	0265	173.0	14.00	165.5	14.20	6.648	86	0.53	028.42	0
028	Manganese (ppm)	0682	173.7	0.0000	165.5	14.20	6.648	86	0.58	028.43	0
028	Manganese (ppm)	0032	173.9	0.7000	165.5	14.20	6.648	86	0.59	028.43	0
028	Manganese (ppm)	2192	174.1	4.830	165.5	14.20	6.648	86	0.61	028.53	0
028	Manganese (ppm)	0003	174.5	1.000	165.5	14.20	6.648	86	0.64	028.41	0
028	Manganese (ppm)	0357	175.1	3.327	165.5	14.20	6.648	86	0.68	028.42	0
028	Manganese (ppm)	2062	175.5	4.600	165.5	14.20	6.648	86	0.71	028.31	0
028	Manganese (ppm)	0027	175.7	1.216	165.5	14.20	6.648	86	0.72	028.43	0
028	Manganese (ppm)	0510	176.0	0.0000	165.5	14.20	6.648	86	0.74	028.43	0
028	Manganese (ppm)	0148	177.0	8.000	165.5	14.20	6.648	86	0.81	028.41	0
028	Manganese (ppm)	0520	177.5	15.00	165.5	14.20	6.648	86	0.85	028.41	0
028	Manganese (ppm)	0038	177.5	11.00	165.5	14.20	6.648	86	0.85	028.43	0
028	Manganese (ppm)	0175	179.0	2.000	165.5	14.20	6.648	86	0.95	028.31	0
028	Manganese (ppm)	0590	179.5	3.000	165.5	14.20	6.648	86	0.99	028.99	0
028	Manganese (ppm)	0726	182.4	0.0000	165.5	14.20	6.648	86	1.19	028.42	0
028	Manganese (ppm)	0100	182.5	5.000	165.5	14.20	6.648	86	1.20	028.99	0
028	Manganese (ppm)	0202	184.0	23.53	165.5	14.20	6.648	86	1.30	028.43	0
028	Manganese (ppm)	0035	185.5	1.000	165.5	14.20	6.648	86	1.41	028.43	0
028	Manganese (ppm)	0035	186.0	8.000	165.5	14.20	6.648	86	1.45	028.42	0
028	Manganese (ppm)	0968	188.5	1.000	165.5	14.20	6.648	86	1.62	028.43	0
028	Manganese (ppm)	0964	189.7	21.53	165.5	14.20	6.648	86	1.70	028.43	0
028	Manganese (ppm)	0190	191.1	3.930	165.5	14.20	6.648	86	1.81	028.42	0
028	Manganese (ppm)	0047	191.4	6.710	165.5	14.20	6.648	86	1.83	028.52	0
028	Manganese (ppm)	0037	202.5	1.800	165.5	14.20	6.648	86	2.61	028.42	0
028	Manganese (ppm)	0536	222.3	20.80	165.5	14.20	6.648	86	4.00	028.31	0
028	Manganese (ppm)	0651	436.1	124.2	165.5	14.20	6.648	86	19.06	028.41	2

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	0405	0.4015	0.0030	0.4899	0.0271	0.0128	117	-3.26	031.41	0
031	Phosphorus (%)	2192	0.4200	0.0400	0.4899	0.0271	0.0128	117	-2.58	031.01	0
031	Phosphorus (%)	0169	0.4250	0.0100	0.4899	0.0271	0.0128	117	-2.40	031.03	0
031	Phosphorus (%)	0032	0.4306	0.0488	0.4899	0.0271	0.0128	117	-2.19	031.43	0
031	Phosphorus (%)	0685	0.4350	0.0100	0.4899	0.0271	0.0128	117	-2.03	031.41	0
031	Phosphorus (%)	0560	0.4420	0.0400	0.4899	0.0271	0.0128	117	-1.77	031.52	0
031	Phosphorus (%)	0018	0.4475	0.0110	0.4899	0.0271	0.0128	117	-1.57	031.01	0
031	Phosphorus (%)	0298	0.4500	0.0200	0.4899	0.0271	0.0128	117	-1.47	031.41	0
031	Phosphorus (%)	0848	0.4500	0.0000	0.4899	0.0271	0.0128	117	-1.47	031.41	0
031	Phosphorus (%)	0186	0.4545	0.0070	0.4899	0.0271	0.0128	117	-1.31	031.52	0
031	Phosphorus (%)	0004	0.4550	0.0500	0.4899	0.0271	0.0128	117	-1.29	031.41	0
031	Phosphorus (%)	0939	0.4550	0.0100	0.4899	0.0271	0.0128	117	-1.29	031.01	0
031	Phosphorus (%)	0353	0.4550	0.0100	0.4899	0.0271	0.0128	117	-1.29	031.43	0
031	Phosphorus (%)	0511	0.4600	0.0200	0.4899	0.0271	0.0128	117	-1.11	031.01	0
031	Phosphorus (%)	0278	0.4600	0.0200	0.4899	0.0271	0.0128	117	-1.11	031.42	0
031	Phosphorus (%)	0910	0.4600	0.0000	0.4899	0.0271	0.0128	117	-1.11	031.41	0
031	Phosphorus (%)	0505	0.4600	0.0000	0.4899	0.0271	0.0128	117	-1.11	031.43	0
031	Phosphorus (%)	0202	0.4630	0.0120	0.4899	0.0271	0.0128	117	-0.99	031.43	0
031	Phosphorus (%)	0100	0.4650	0.0100	0.4899	0.0271	0.0128	117	-0.92	031.99	0
031	Phosphorus (%)	0208	0.4680	0.0040	0.4899	0.0271	0.0128	117	-0.81	031.01	0
031	Phosphorus (%)	0186	0.4695	0.0090	0.4899	0.0271	0.0128	117	-0.75	031.42	0
031	Phosphorus (%)	0175	0.4700	0.0200	0.4899	0.0271	0.0128	117	-0.74	031.01	0
031	Phosphorus (%)	0354	0.4700	0.0000	0.4899	0.0271	0.0128	117	-0.74	031.41	0
031	Phosphorus (%)	0010	0.4700	0.0000	0.4899	0.0271	0.0128	117	-0.74	031.43	0
031	Phosphorus (%)	0297	0.4700	0.0200	0.4899	0.0271	0.0128	117	-0.74	031.43	0
031	Phosphorus (%)	2109	0.4735	0.0130	0.4899	0.0271	0.0128	117	-0.61	031.01	0
031	Phosphorus (%)	0164	0.4745	0.0010	0.4899	0.0271	0.0128	117	-0.57	031.44	0
031	Phosphorus (%)	0878	0.4750	0.0300	0.4899	0.0271	0.0128	117	-0.55	031.01	0
031	Phosphorus (%)	0897	0.4750	0.0100	0.4899	0.0271	0.0128	117	-0.55	031.01	0
031	Phosphorus (%)	2181	0.4750	0.0100	0.4899	0.0271	0.0128	117	-0.55	031.01	0
031	Phosphorus (%)	0049	0.4750	0.0100	0.4899	0.0271	0.0128	117	-0.55	031.41	0
031	Phosphorus (%)	0123	0.4750	0.0100	0.4899	0.0271	0.0128	117	-0.55	031.41	0
031	Phosphorus (%)	2259	0.4750	0.0040	0.4899	0.0271	0.0128	117	-0.55	031.41	0
031	Phosphorus (%)	0229	0.4760	0.0120	0.4899	0.0271	0.0128	117	-0.51	031.41	0
031	Phosphorus (%)	0139	0.4765	0.0070	0.4899	0.0271	0.0128	117	-0.50	031.41	0
031	Phosphorus (%)	0407	0.4772	0.0071	0.4899	0.0271	0.0128	117	-0.47	031.41	0
031	Phosphorus (%)	0345	0.4775	0.0050	0.4899	0.0271	0.0128	117	-0.46	031.43	0
031	Phosphorus (%)	0653	0.4780	0.0020	0.4899	0.0271	0.0128	117	-0.44	031.41	0
031	Phosphorus (%)	0622	0.4781	0.0089	0.4899	0.0271	0.0128	117	-0.44	031.01	0
031	Phosphorus (%)	0142	0.4792	0.0094	0.4899	0.0271	0.0128	117	-0.40	031.01	0
031	Phosphorus (%)	0529	0.4800	0.0200	0.4899	0.0271	0.0128	117	-0.37	031.01	0
031	Phosphorus (%)	0590	0.4800	0.0000	0.4899	0.0271	0.0128	117	-0.37	031.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	2188	0.4800	0.0000	0.4899	0.0271	0.0128	117	-0.37	031.01	0
031	Phosphorus (%)	0171	0.4800	0.0000	0.4899	0.0271	0.0128	117	-0.37	031.41	0
031	Phosphorus (%)	0265	0.4800	0.0200	0.4899	0.0271	0.0128	117	-0.37	031.42	0
031	Phosphorus (%)	0016	0.4820	0.0160	0.4899	0.0271	0.0128	117	-0.29	031.01	0
031	Phosphorus (%)	2196	0.4820	0.0000	0.4899	0.0271	0.0128	117	-0.29	031.01	0
031	Phosphorus (%)	0036	0.4824	0.0057	0.4899	0.0271	0.0128	117	-0.28	031.03	0
031	Phosphorus (%)	0001	0.4845	0.0091	0.4899	0.0271	0.0128	117	-0.20	031.03	0
031	Phosphorus (%)	0563	0.4848	0.0065	0.4899	0.0271	0.0128	117	-0.19	031.01	0
031	Phosphorus (%)	0689	0.4850	0.0100	0.4899	0.0271	0.0128	117	-0.18	031.01	0
031	Phosphorus (%)	0019	0.4850	0.0100	0.4899	0.0271	0.0128	117	-0.18	031.41	0
031	Phosphorus (%)	0366	0.4850	0.0300	0.4899	0.0271	0.0128	117	-0.18	031.42	0
031	Phosphorus (%)	0610	0.4850	0.0100	0.4899	0.0271	0.0128	117	-0.18	031.43	0
031	Phosphorus (%)	0939	0.4850	0.0300	0.4899	0.0271	0.0128	117	-0.18	031.53	0
031	Phosphorus (%)	0242	0.4850	0.0100	0.4899	0.0271	0.0128	117	-0.18	031.99	0
031	Phosphorus (%)	0001	0.4865	0.0170	0.4899	0.0271	0.0128	117	-0.13	031.43	0
031	Phosphorus (%)	0651	0.4866	0.0067	0.4899	0.0271	0.0128	117	-0.13	031.01	0
031	Phosphorus (%)	0208	0.4880	0.0060	0.4899	0.0271	0.0128	117	-0.07	031.41	0
031	Phosphorus (%)	0693	0.4890	0.0260	0.4899	0.0271	0.0128	117	-0.04	031.42	0
031	Phosphorus (%)	0357	0.4892	0.0022	0.4899	0.0271	0.0128	117	-0.03	031.42	0
031	Phosphorus (%)	0083	0.4900	0.0000	0.4899	0.0271	0.0128	117	0.00	031.43	0
031	Phosphorus (%)	0870	0.4902	0.0013	0.4899	0.0271	0.0128	117	0.01	031.42	0
031	Phosphorus (%)	0017	0.4905	0.0003	0.4899	0.0271	0.0128	117	0.02	031.43	0
031	Phosphorus (%)	0037	0.4905	0.0070	0.4899	0.0271	0.0128	117	0.02	031.42	0
031	Phosphorus (%)	0948	0.4910	0.0020	0.4899	0.0271	0.0128	117	0.04	031.01	0
031	Phosphorus (%)	0407	0.4933	0.0351	0.4899	0.0271	0.0128	117	0.12	031.43	0
031	Phosphorus (%)	0881	0.4935	0.0190	0.4899	0.0271	0.0128	117	0.13	031.01	0
031	Phosphorus (%)	0626	0.4950	0.0420	0.4899	0.0271	0.0128	117	0.19	031.01	0
031	Phosphorus (%)	2199	0.4950	0.0100	0.4899	0.0271	0.0128	117	0.19	031.01	0
031	Phosphorus (%)	0144	0.4950	0.0100	0.4899	0.0271	0.0128	117	0.19	031.41	0
031	Phosphorus (%)	0675	0.4950	0.0100	0.4899	0.0271	0.0128	117	0.19	031.43	0
031	Phosphorus (%)	0676	0.4950	0.0300	0.4899	0.0271	0.0128	117	0.19	031.99	0
031	Phosphorus (%)	0968	0.4955	0.0010	0.4899	0.0271	0.0128	117	0.20	031.43	0
031	Phosphorus (%)	0027	0.4965	0.0110	0.4899	0.0271	0.0128	117	0.24	031.43	0
031	Phosphorus (%)	0043	0.4968	0.0052	0.4899	0.0271	0.0128	117	0.25	031.02	0
031	Phosphorus (%)	0598	0.4969	0.0022	0.4899	0.0271	0.0128	117	0.26	031.41	0
031	Phosphorus (%)	0006	0.4985	0.0170	0.4899	0.0271	0.0128	117	0.32	031.43	0
031	Phosphorus (%)	2129	0.4990	0.0014	0.4899	0.0271	0.0128	117	0.33	031.42	0
031	Phosphorus (%)	2009	0.4991	0.0008	0.4899	0.0271	0.0128	117	0.34	031.01	0
031	Phosphorus (%)	0036	0.4992	0.0114	0.4899	0.0271	0.0128	117	0.34	031.43	0
031	Phosphorus (%)	0674	0.5000	0.0000	0.4899	0.0271	0.0128	117	0.37	031.01	0
031	Phosphorus (%)	0148	0.5000	0.0000	0.4899	0.0271	0.0128	117	0.37	031.41	0
031	Phosphorus (%)	0726	0.5000	0.0000	0.4899	0.0271	0.0128	117	0.37	031.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	0682	0.5000	0.0000	0.4899	0.0271	0.0128	117	0.37	031.43	0
031	Phosphorus (%)	0045	0.5020	0.0020	0.4899	0.0271	0.0128	117	0.44	031.42	0
031	Phosphorus (%)	0668	0.5025	0.0030	0.4899	0.0271	0.0128	117	0.46	031.43	0
031	Phosphorus (%)	0014	0.5035	0.0150	0.4899	0.0271	0.0128	117	0.50	031.42	0
031	Phosphorus (%)	2193	0.5050	0.0100	0.4899	0.0271	0.0128	117	0.56	031.01	0
031	Phosphorus (%)	0190	0.5050	0.0100	0.4899	0.0271	0.0128	117	0.56	031.42	0
031	Phosphorus (%)	0035	0.5100	0.0200	0.4899	0.0271	0.0128	117	0.74	031.42	0
031	Phosphorus (%)	0038	0.5125	0.0150	0.4899	0.0271	0.0128	117	0.83	031.43	0
031	Phosphorus (%)	0512	0.5127	0.0122	0.4899	0.0271	0.0128	117	0.84	031.43	0
031	Phosphorus (%)	0918	0.5135	0.0070	0.4899	0.0271	0.0128	117	0.87	031.43	0
031	Phosphorus (%)	0042	0.5145	0.0070	0.4899	0.0271	0.0128	117	0.91	031.42	0
031	Phosphorus (%)	0160	0.5150	0.0100	0.4899	0.0271	0.0128	117	0.92	031.44	0
031	Phosphorus (%)	2290	0.5187	0.0509	0.4899	0.0271	0.0128	117	1.06	031.99	0
031	Phosphorus (%)	0074	0.5190	0.0000	0.4899	0.0271	0.0128	117	1.07	031.41	0
031	Phosphorus (%)	0098	0.5192	0.0192	0.4899	0.0271	0.0128	117	1.08	031.43	0
031	Phosphorus (%)	0226	0.5200	0.0200	0.4899	0.0271	0.0128	117	1.11	031.41	0
031	Phosphorus (%)	0510	0.5200	0.0000	0.4899	0.0271	0.0128	117	1.11	031.43	0
031	Phosphorus (%)	0692	0.5250	0.0100	0.4899	0.0271	0.0128	117	1.29	031.42	0
031	Phosphorus (%)	0043	0.5276	0.0200	0.4899	0.0271	0.0128	117	1.39	031.43	0
031	Phosphorus (%)	0033	0.5300	0.0400	0.4899	0.0271	0.0128	117	1.48	031.43	0
031	Phosphorus (%)	0964	0.5320	0.0575	0.4899	0.0271	0.0128	117	1.55	031.43	0
031	Phosphorus (%)	0098	0.5339	0.0254	0.4899	0.0271	0.0128	117	1.62	031.41	0
031	Phosphorus (%)	2006	0.5345	0.0030	0.4899	0.0271	0.0128	117	1.64	031.01	0
031	Phosphorus (%)	0358	0.5350	0.0300	0.4899	0.0271	0.0128	117	1.66	031.41	0
031	Phosphorus (%)	0035	0.5400	0.0000	0.4899	0.0271	0.0128	117	1.85	031.43	0
031	Phosphorus (%)	0619	0.5485	0.0090	0.4899	0.0271	0.0128	117	2.16	031.41	0
031	Phosphorus (%)	0520	0.5600	0.0000	0.4899	0.0271	0.0128	117	2.59	031.41	0
031	Phosphorus (%)	0852	0.5600	0.0200	0.4899	0.0271	0.0128	117	2.59	031.99	0
031	Phosphorus (%)	0263	0.5646	0.0028	0.4899	0.0271	0.0128	117	2.76	031.42	0
031	Phosphorus (%)	0536	0.5700	0.0000	0.4899	0.0271	0.0128	117	2.95	031.06	0
031	Phosphorus (%)	0003	0.5700	0.0600	0.4899	0.0271	0.0128	117	2.95	031.41	0
031	Phosphorus (%)	0047	0.7000	0.0400	0.4899	0.0271	0.0128	117	7.75	031.42	0
031	Phosphorus (%)	2062	0.9081	0.0079	0.4899	0.0271	0.0128	117	15.43	031.01	0
031	Phosphorus (%)	0687	0.4850	0.0700	0.4899	0.0271	0.0128	117	-0.18	031.01	1
031	Phosphorus (%)	2146	1.155	0.0300	0.4899	0.0271	0.0128	117	24.55	031.01	2
031	Phosphorus (%)	2146	1.195	0.0500	0.4899	0.0271	0.0128	117	26.02	031.51	2
032	Potassium (%)	0536	0.5350	0.0300	0.8464	0.0461	0.0206	83	-6.76	032.31	0
032	Potassium (%)	0003	0.7150	0.0700	0.8464	0.0461	0.0206	83	-2.85	032.41	0
032	Potassium (%)	0598	0.7410	0.0316	0.8464	0.0461	0.0206	83	-2.29	032.41	0
032	Potassium (%)	0511	0.7450	0.0100	0.8464	0.0461	0.0206	83	-2.20	032.41	0
032	Potassium (%)	0049	0.7550	0.0100	0.8464	0.0461	0.0206	83	-1.98	032.41	0
032	Potassium (%)	0354	0.7600	0.0000	0.8464	0.0461	0.0206	83	-1.87	032.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
032	Potassium (%)	0032	0.7691	0.0101	0.8464	0.0461	0.0206	83	-1.68	032.43	0
032	Potassium (%)	0175	0.7700	0.0600	0.8464	0.0461	0.0206	83	-1.66	032.31	0
032	Potassium (%)	0004	0.7700	0.0400	0.8464	0.0461	0.0206	83	-1.66	032.41	0
032	Potassium (%)	0186	0.7925	0.0130	0.8464	0.0461	0.0206	83	-1.17	032.52	0
032	Potassium (%)	0019	0.7950	0.0100	0.8464	0.0461	0.0206	83	-1.12	032.41	0
032	Potassium (%)	0265	0.7950	0.0100	0.8464	0.0461	0.0206	83	-1.12	032.42	0
032	Potassium (%)	0405	0.8005	0.0190	0.8464	0.0461	0.0206	83	-1.00	032.41	0
032	Potassium (%)	0563	0.8009	0.0119	0.8464	0.0461	0.0206	83	-0.99	032.31	0
032	Potassium (%)	0229	0.8025	0.0010	0.8464	0.0461	0.0206	83	-0.95	032.41	0
032	Potassium (%)	0278	0.8050	0.0100	0.8464	0.0461	0.0206	83	-0.90	032.42	0
032	Potassium (%)	0038	0.8050	0.0100	0.8464	0.0461	0.0206	83	-0.90	032.43	0
032	Potassium (%)	0939	0.8100	0.0000	0.8464	0.0461	0.0206	83	-0.79	032.31	0
032	Potassium (%)	0870	0.8139	0.0412	0.8464	0.0461	0.0206	83	-0.71	032.42	0
032	Potassium (%)	2146	0.8150	0.0100	0.8464	0.0461	0.0206	83	-0.68	032.31	0
032	Potassium (%)	0144	0.8150	0.0300	0.8464	0.0461	0.0206	83	-0.68	032.41	0
032	Potassium (%)	0345	0.8175	0.0090	0.8464	0.0461	0.0206	83	-0.63	032.43	0
032	Potassium (%)	0186	0.8190	0.0020	0.8464	0.0461	0.0206	83	-0.59	032.42	0
032	Potassium (%)	0910	0.8200	0.0400	0.8464	0.0461	0.0206	83	-0.57	032.41	0
032	Potassium (%)	0208	0.8220	0.0040	0.8464	0.0461	0.0206	83	-0.53	032.41	0
032	Potassium (%)	2129	0.8235	0.0339	0.8464	0.0461	0.0206	83	-0.50	032.42	0
032	Potassium (%)	0001	0.8270	0.0120	0.8464	0.0461	0.0206	83	-0.42	032.99	0
032	Potassium (%)	0675	0.8300	0.0200	0.8464	0.0461	0.0206	83	-0.36	032.43	0
032	Potassium (%)	0074	0.8345	0.0130	0.8464	0.0461	0.0206	83	-0.26	032.41	0
032	Potassium (%)	0619	0.8355	0.0130	0.8464	0.0461	0.0206	83	-0.24	032.41	0
032	Potassium (%)	0505	0.8400	0.0400	0.8464	0.0461	0.0206	83	-0.14	032.43	0
032	Potassium (%)	2192	0.8400	0.0400	0.8464	0.0461	0.0206	83	-0.14	032.43	0
032	Potassium (%)	0918	0.8410	0.0080	0.8464	0.0461	0.0206	83	-0.12	032.43	0
032	Potassium (%)	0142	0.8417	0.0087	0.8464	0.0461	0.0206	83	-0.10	032.31	0
032	Potassium (%)	0045	0.8430	0.0340	0.8464	0.0461	0.0206	83	-0.07	032.42	0
032	Potassium (%)	0139	0.8450	0.0100	0.8464	0.0461	0.0206	83	-0.03	032.41	0
032	Potassium (%)	0148	0.8450	0.0900	0.8464	0.0461	0.0206	83	-0.03	032.41	0
032	Potassium (%)	0033	0.8450	0.0100	0.8464	0.0461	0.0206	83	-0.03	032.43	0
032	Potassium (%)	0035	0.8450	0.0100	0.8464	0.0461	0.0206	83	-0.03	032.43	0
032	Potassium (%)	0353	0.8450	0.0100	0.8464	0.0461	0.0206	83	-0.03	032.43	0
032	Potassium (%)	0939	0.8450	0.0100	0.8464	0.0461	0.0206	83	-0.03	032.53	0
032	Potassium (%)	0560	0.8477	0.0669	0.8464	0.0461	0.0206	83	0.03	032.52	0
032	Potassium (%)	0529	0.8500	0.0200	0.8464	0.0461	0.0206	83	0.08	032.31	0
032	Potassium (%)	0035	0.8500	0.0000	0.8464	0.0461	0.0206	83	0.08	032.42	0
032	Potassium (%)	0357	0.8509	0.0013	0.8464	0.0461	0.0206	83	0.10	032.42	0
032	Potassium (%)	0027	0.8520	0.0080	0.8464	0.0461	0.0206	83	0.12	032.43	0
032	Potassium (%)	0512	0.8535	0.0067	0.8464	0.0461	0.0206	83	0.15	032.43	0
032	Potassium (%)	0878	0.8550	0.0100	0.8464	0.0461	0.0206	83	0.19	032.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
032	Potassium (%)	0098	0.8550	0.0900	0.8464	0.0461	0.0206	83	0.19	032.43	0
032	Potassium (%)	0100	0.8550	0.0300	0.8464	0.0461	0.0206	83	0.19	032.99	0
032	Potassium (%)	0668	0.8560	0.0160	0.8464	0.0461	0.0206	83	0.21	032.43	0
032	Potassium (%)	0968	0.8570	0.0020	0.8464	0.0461	0.0206	83	0.23	032.43	0
032	Potassium (%)	0682	0.8600	0.0000	0.8464	0.0461	0.0206	83	0.29	032.43	0
032	Potassium (%)	0948	0.8610	0.0220	0.8464	0.0461	0.0206	83	0.32	032.31	0
032	Potassium (%)	0047	0.8650	0.0100	0.8464	0.0461	0.0206	83	0.40	032.42	0
032	Potassium (%)	0010	0.8650	0.0100	0.8464	0.0461	0.0206	83	0.40	032.43	0
032	Potassium (%)	0653	0.8660	0.0020	0.8464	0.0461	0.0206	83	0.42	032.41	0
032	Potassium (%)	0693	0.8685	0.0350	0.8464	0.0461	0.0206	83	0.48	032.42	0
032	Potassium (%)	0358	0.8700	0.0000	0.8464	0.0461	0.0206	83	0.51	032.41	0
032	Potassium (%)	0520	0.8700	0.0200	0.8464	0.0461	0.0206	83	0.51	032.41	0
032	Potassium (%)	0692	0.8700	0.0400	0.8464	0.0461	0.0206	83	0.51	032.42	0
032	Potassium (%)	0726	0.8700	0.0000	0.8464	0.0461	0.0206	83	0.51	032.42	0
032	Potassium (%)	0164	0.8700	0.0020	0.8464	0.0461	0.0206	83	0.51	032.44	0
032	Potassium (%)	0242	0.8700	0.0400	0.8464	0.0461	0.0206	83	0.51	032.99	0
032	Potassium (%)	0208	0.8720	0.0160	0.8464	0.0461	0.0206	83	0.55	032.31	0
032	Potassium (%)	0037	0.8835	0.0010	0.8464	0.0461	0.0206	83	0.80	032.42	0
032	Potassium (%)	0042	0.8845	0.0330	0.8464	0.0461	0.0206	83	0.83	032.42	0
032	Potassium (%)	0083	0.8850	0.0100	0.8464	0.0461	0.0206	83	0.84	032.43	0
032	Potassium (%)	0202	0.8850	0.0160	0.8464	0.0461	0.0206	83	0.84	032.43	0
032	Potassium (%)	0407	0.8859	0.0471	0.8464	0.0461	0.0206	83	0.86	032.43	0
032	Potassium (%)	0160	0.8900	0.0200	0.8464	0.0461	0.0206	83	0.95	032.44	0
032	Potassium (%)	0297	0.9000	0.0200	0.8464	0.0461	0.0206	83	1.16	032.43	0
032	Potassium (%)	0169	0.9050	0.0300	0.8464	0.0461	0.0206	83	1.27	032.32	0
032	Potassium (%)	0190	0.9050	0.0100	0.8464	0.0461	0.0206	83	1.27	032.42	0
032	Potassium (%)	0610	0.9050	0.0100	0.8464	0.0461	0.0206	83	1.27	032.43	0
032	Potassium (%)	0017	0.9064	0.0111	0.8464	0.0461	0.0206	83	1.30	032.43	0
032	Potassium (%)	0226	0.9150	0.0100	0.8464	0.0461	0.0206	83	1.49	032.41	0
032	Potassium (%)	0171	0.9200	0.0600	0.8464	0.0461	0.0206	83	1.60	032.41	0
032	Potassium (%)	0366	0.9200	0.0400	0.8464	0.0461	0.0206	83	1.60	032.42	0
032	Potassium (%)	0510	0.9200	0.0000	0.8464	0.0461	0.0206	83	1.60	032.43	0
032	Potassium (%)	0098	0.9350	0.0500	0.8464	0.0461	0.0206	83	1.92	032.41	0
032	Potassium (%)	0964	0.9435	0.0367	0.8464	0.0461	0.0206	83	2.10	032.43	0
032	Potassium (%)	0263	1.041	0.0156	0.8464	0.0461	0.0206	83	4.22	032.42	0
032	Potassium (%)	0407	0.8223	0.1137	0.8464	0.0461	0.0206	83	-0.52	032.41	1
032	Potassium (%)	2146	0.8750	0.1300	0.8464	0.0461	0.0206	83	0.62	032.51	1
032	Potassium (%)	0123	0.1650	0.0100	0.8464	0.0461	0.0206	83	-14.78	032.41	2
033	Salt as chloride (%)	0003	0.1600	0.0200	0.3636	0.0479	0.0158	62	-4.25	033.99	0
033	Salt as chloride (%)	0693	0.2610	0.0360	0.3636	0.0479	0.0158	62	-2.14	033.00	0
033	Salt as chloride (%)	2073	0.2650	0.0100	0.3636	0.0479	0.0158	62	-2.06	033.99	0
033	Salt as chloride (%)	0144	0.2700	0.0200	0.3636	0.0479	0.0158	62	-1.95	033.03	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
033	Salt as chloride (%)	0265	0.2750	0.0100	0.3636	0.0479	0.0158	62	-1.85	033.03	0
033	Salt as chloride (%)	0505	0.2900	0.0200	0.3636	0.0479	0.0158	62	-1.54	033.03	0
033	Salt as chloride (%)	0619	0.2975	0.0250	0.3636	0.0479	0.0158	62	-1.38	033.01	0
033	Salt as chloride (%)	0171	0.3050	0.0100	0.3636	0.0479	0.0158	62	-1.22	033.00	0
033	Salt as chloride (%)	0682	0.3100	0.0000	0.3636	0.0479	0.0158	62	-1.12	033.99	0
033	Salt as chloride (%)	0190	0.3100	0.0600	0.3636	0.0479	0.0158	62	-1.12	033.03	0
033	Salt as chloride (%)	2181	0.3100	0.0600	0.3636	0.0479	0.0158	62	-1.12	033.99	0
033	Salt as chloride (%)	0726	0.3200	0.0000	0.3636	0.0479	0.0158	62	-0.91	033.03	0
033	Salt as chloride (%)	2062	0.3253	0.0031	0.3636	0.0479	0.0158	62	-0.80	033.00	0
033	Salt as chloride (%)	2009	0.3292	0.0044	0.3636	0.0479	0.0158	62	-0.72	033.00	0
033	Salt as chloride (%)	0878	0.3300	0.0200	0.3636	0.0479	0.0158	62	-0.70	033.00	0
033	Salt as chloride (%)	0083	0.3300	0.0000	0.3636	0.0479	0.0158	62	-0.70	033.01	0
033	Salt as chloride (%)	0897	0.3350	0.0100	0.3636	0.0479	0.0158	62	-0.60	033.00	0
033	Salt as chloride (%)	0650	0.3450	0.0100	0.3636	0.0479	0.0158	62	-0.39	033.99	0
033	Salt as chloride (%)	0123	0.3500	0.0200	0.3636	0.0479	0.0158	62	-0.28	033.00	0
033	Salt as chloride (%)	0921	0.3500	0.0000	0.3636	0.0479	0.0158	62	-0.28	033.00	0
033	Salt as chloride (%)	0910	0.3500	0.0200	0.3636	0.0479	0.0158	62	-0.28	033.99	0
033	Salt as chloride (%)	0309	0.3527	0.0034	0.3636	0.0479	0.0158	62	-0.23	033.00	0
033	Salt as chloride (%)	2129	0.3535	0.0090	0.3636	0.0479	0.0158	62	-0.21	033.99	0
033	Salt as chloride (%)	0208	0.3545	0.0130	0.3636	0.0479	0.0158	62	-0.19	033.99	0
033	Salt as chloride (%)	0169	0.3550	0.0100	0.3636	0.0479	0.0158	62	-0.18	033.00	0
033	Salt as chloride (%)	0001	0.3550	0.0100	0.3636	0.0479	0.0158	62	-0.18	033.01	0
033	Salt as chloride (%)	0148	0.3550	0.0100	0.3636	0.0479	0.0158	62	-0.18	033.01	0
033	Salt as chloride (%)	0354	0.3550	0.0100	0.3636	0.0479	0.0158	62	-0.18	033.01	0
033	Salt as chloride (%)	2076	0.3600	0.0600	0.3636	0.0479	0.0158	62	-0.07	033.00	0
033	Salt as chloride (%)	0242	0.3600	0.0000	0.3636	0.0479	0.0158	62	-0.07	033.01	0
033	Salt as chloride (%)	0407	0.3625	0.0050	0.3636	0.0479	0.0158	62	-0.02	033.01	0
033	Salt as chloride (%)	0004	0.3650	0.0100	0.3636	0.0479	0.0158	62	0.03	033.01	0
033	Salt as chloride (%)	0689	0.3650	0.0100	0.3636	0.0479	0.0158	62	0.03	033.05	0
033	Salt as chloride (%)	2196	0.3700	0.0000	0.3636	0.0479	0.0158	62	0.13	033.00	0
033	Salt as chloride (%)	0939	0.3700	0.0000	0.3636	0.0479	0.0158	62	0.13	033.01	0
033	Salt as chloride (%)	0510	0.3710	0.0160	0.3636	0.0479	0.0158	62	0.15	033.01	0
033	Salt as chloride (%)	2109	0.3725	0.0070	0.3636	0.0479	0.0158	62	0.19	033.01	0
033	Salt as chloride (%)	0175	0.3750	0.0100	0.3636	0.0479	0.0158	62	0.24	033.01	0
033	Salt as chloride (%)	0366	0.3800	0.0200	0.3636	0.0479	0.0158	62	0.34	033.00	0
033	Salt as chloride (%)	0164	0.3800	0.0000	0.3636	0.0479	0.0158	62	0.34	033.01	0
033	Salt as chloride (%)	0653	0.3835	0.0150	0.3636	0.0479	0.0158	62	0.42	033.00	0
033	Salt as chloride (%)	0353	0.3900	0.0600	0.3636	0.0479	0.0158	62	0.55	033.00	0
033	Salt as chloride (%)	0881	0.3900	0.0120	0.3636	0.0479	0.0158	62	0.55	033.00	0
033	Salt as chloride (%)	0100	0.3900	0.0200	0.3636	0.0479	0.0158	62	0.55	033.01	0
033	Salt as chloride (%)	0226	0.3900	0.0200	0.3636	0.0479	0.0158	62	0.55	033.01	0
033	Salt as chloride (%)	0032	0.3900	0.0000	0.3636	0.0479	0.0158	62	0.55	033.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
033	Salt as chloride (%)	0045	0.3915	0.0110	0.3636	0.0479	0.0158	62	0.58	033.00	0
033	Salt as chloride (%)	0511	0.3950	0.0300	0.3636	0.0479	0.0158	62	0.66	033.00	0
033	Salt as chloride (%)	0590	0.3950	0.0100	0.3636	0.0479	0.0158	62	0.66	033.01	0
033	Salt as chloride (%)	0003	0.4000	0.0000	0.3636	0.0479	0.0158	62	0.76	033.01	0
033	Salt as chloride (%)	0229	0.4000	0.0000	0.3636	0.0479	0.0158	62	0.76	033.01	0
033	Salt as chloride (%)	0675	0.4100	0.0400	0.3636	0.0479	0.0158	62	0.97	033.00	0
033	Salt as chloride (%)	0948	0.4100	0.0200	0.3636	0.0479	0.0158	62	0.97	033.01	0
033	Salt as chloride (%)	0017	0.4111	0.0061	0.3636	0.0479	0.0158	62	0.99	033.99	0
033	Salt as chloride (%)	0536	0.4150	0.0100	0.3636	0.0479	0.0158	62	1.07	033.99	0
033	Salt as chloride (%)	0019	0.4150	0.0300	0.3636	0.0479	0.0158	62	1.07	033.01	0
033	Salt as chloride (%)	0098	0.4300	0.0200	0.3636	0.0479	0.0158	62	1.39	033.01	0
033	Salt as chloride (%)	2193	0.4500	0.0200	0.3636	0.0479	0.0158	62	1.80	033.00	0
033	Salt as chloride (%)	0298	0.4600	0.0200	0.3636	0.0479	0.0158	62	2.01	033.00	0
033	Salt as chloride (%)	0358	0.4750	0.0300	0.3636	0.0479	0.0158	62	2.33	033.99	0
033	Salt as chloride (%)	0539	0.4895	0.0150	0.3636	0.0479	0.0158	62	2.63	033.00	0
033	Salt as chloride (%)	0278	0.7250	0.0300	0.3636	0.0479	0.0158	62	7.55	033.01	0
033	Salt as chloride (%)	2146	2.115	0.0100	0.3636	0.0479	0.0158	62	36.58	033.01	2
034	Selenium (ppm)	0047	0.4050	0.0700	0.5859	0.1593	0.0610	22	-1.14	034.52	0
034	Selenium (ppm)	0964	0.4066	0.0180	0.5859	0.1593	0.0610	22	-1.13	034.43	0
034	Selenium (ppm)	0171	0.4170	0.0324	0.5859	0.1593	0.0610	22	-1.06	034.04	0
034	Selenium (ppm)	0619	0.4255	0.0310	0.5859	0.1593	0.0610	22	-1.01	034.41	0
034	Selenium (ppm)	0045	0.4505	0.0130	0.5859	0.1593	0.0610	22	-0.85	034.04	0
034	Selenium (ppm)	0939	0.4650	0.0500	0.5859	0.1593	0.0610	22	-0.76	034.53	0
034	Selenium (ppm)	0169	0.5050	0.0100	0.5859	0.1593	0.0610	22	-0.51	034.04	0
034	Selenium (ppm)	0563	0.5213	0.0525	0.5859	0.1593	0.0610	22	-0.41	034.04	0
034	Selenium (ppm)	0164	0.5310	0.0400	0.5859	0.1593	0.0610	22	-0.34	034.53	0
034	Selenium (ppm)	0148	0.5400	0.0000	0.5859	0.1593	0.0610	22	-0.29	034.41	0
034	Selenium (ppm)	0038	0.5862	0.0234	0.5859	0.1593	0.0610	22	0.00	034.01	0
034	Selenium (ppm)	0910	0.5900	0.0400	0.5859	0.1593	0.0610	22	0.03	034.52	0
034	Selenium (ppm)	0003	0.5900	0.1720	0.5859	0.1593	0.0610	22	0.03	034.53	0
034	Selenium (ppm)	0227	0.6030	0.0100	0.5859	0.1593	0.0610	22	0.11	034.53	0
034	Selenium (ppm)	0918	0.6109	0.0059	0.5859	0.1593	0.0610	22	0.16	034.53	0
034	Selenium (ppm)	0208	0.6145	0.0570	0.5859	0.1593	0.0610	22	0.18	034.52	0
034	Selenium (ppm)	0560	0.6185	0.0770	0.5859	0.1593	0.0610	22	0.20	034.53	0
034	Selenium (ppm)	0032	0.7100	0.1420	0.5859	0.1593	0.0610	22	0.78	034.53	0
034	Selenium (ppm)	0968	0.9280	0.0280	0.5859	0.1593	0.0610	22	2.15	034.43	0
034	Selenium (ppm)	0692	1.150	0.1000	0.5859	0.1593	0.0610	22	3.54	034.42	0
034	Selenium (ppm)	0407	1.390	0.1600	0.5859	0.1593	0.0610	22	5.05	034.41	0
034	Selenium (ppm)	0190	3.835	0.2100	0.5859	0.1593	0.0610	22	20.39	034.99	0
034	Selenium (ppm)	0407	1.775	0.8700	0.5859	0.1593	0.0610	22	7.46	034.43	1
034	Selenium (ppm)	2302	< 0.1		0.5859	0.1593	0.0610	22		034.99	5
034	Selenium (ppm)	0682	< 0.5		0.5859	0.1593	0.0610	22		034.43	5

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
035	Sodium (%)	0175	0.1250	0.0100	0.1692	0.0138	0.0082	86	-3.21	035.31	0
035	Sodium (%)	0536	0.1350	0.0100	0.1692	0.0138	0.0082	86	-2.48	035.31	0
035	Sodium (%)	0511	0.1400	0.0000	0.1692	0.0138	0.0082	86	-2.12	035.41	0
035	Sodium (%)	0298	0.1500	0.0200	0.1692	0.0138	0.0082	86	-1.40	035.41	0
035	Sodium (%)	2129	0.1503	0.0057	0.1692	0.0138	0.0082	86	-1.38	035.42	0
035	Sodium (%)	0345	0.1510	0.0020	0.1692	0.0138	0.0082	86	-1.32	035.43	0
035	Sodium (%)	0529	0.1525	0.0010	0.1692	0.0138	0.0082	86	-1.21	035.31	0
035	Sodium (%)	0353	0.1537	0.0017	0.1692	0.0138	0.0082	86	-1.13	035.43	0
035	Sodium (%)	0144	0.1550	0.0100	0.1692	0.0138	0.0082	86	-1.03	035.41	0
035	Sodium (%)	0265	0.1550	0.0100	0.1692	0.0138	0.0082	86	-1.03	035.42	0
035	Sodium (%)	0098	0.1550	0.0100	0.1692	0.0138	0.0082	86	-1.03	035.43	0
035	Sodium (%)	0939	0.1550	0.0100	0.1692	0.0138	0.0082	86	-1.03	035.53	0
035	Sodium (%)	0870	0.1555	0.0071	0.1692	0.0138	0.0082	86	-1.00	035.42	0
035	Sodium (%)	0512	0.1568	0.0083	0.1692	0.0138	0.0082	86	-0.91	035.43	0
035	Sodium (%)	0405	0.1570	0.0040	0.1692	0.0138	0.0082	86	-0.89	035.41	0
035	Sodium (%)	0045	0.1570	0.0020	0.1692	0.0138	0.0082	86	-0.89	035.42	0
035	Sodium (%)	0560	0.1589	0.0061	0.1692	0.0138	0.0082	86	-0.75	035.52	0
035	Sodium (%)	0598	0.1599	0.0027	0.1692	0.0138	0.0082	86	-0.68	035.41	0
035	Sodium (%)	0939	0.1600	0.0000	0.1692	0.0138	0.0082	86	-0.67	035.31	0
035	Sodium (%)	0171	0.1600	0.0000	0.1692	0.0138	0.0082	86	-0.67	035.41	0
035	Sodium (%)	0035	0.1600	0.0000	0.1692	0.0138	0.0082	86	-0.67	035.42	0
035	Sodium (%)	0037	0.1600	0.0000	0.1692	0.0138	0.0082	86	-0.67	035.42	0
035	Sodium (%)	0033	0.1600	0.0000	0.1692	0.0138	0.0082	86	-0.67	035.43	0
035	Sodium (%)	0035	0.1600	0.0000	0.1692	0.0138	0.0082	86	-0.67	035.43	0
035	Sodium (%)	0563	0.1617	0.0010	0.1692	0.0138	0.0082	86	-0.55	035.31	0
035	Sodium (%)	0017	0.1617	0.0024	0.1692	0.0138	0.0082	86	-0.55	035.43	0
035	Sodium (%)	0407	0.1624	0.0096	0.1692	0.0138	0.0082	86	-0.50	035.43	0
035	Sodium (%)	0918	0.1625	0.0010	0.1692	0.0138	0.0082	86	-0.49	035.43	0
035	Sodium (%)	0881	0.1630	0.0020	0.1692	0.0138	0.0082	86	-0.45	035.43	0
035	Sodium (%)	0043	0.1636	0.0249	0.1692	0.0138	0.0082	86	-0.41	035.43	0
035	Sodium (%)	0407	0.1646	0.0098	0.1692	0.0138	0.0082	86	-0.34	035.41	0
035	Sodium (%)	0019	0.1650	0.0100	0.1692	0.0138	0.0082	86	-0.31	035.41	0
035	Sodium (%)	0123	0.1650	0.0100	0.1692	0.0138	0.0082	86	-0.31	035.41	0
035	Sodium (%)	0354	0.1650	0.0100	0.1692	0.0138	0.0082	86	-0.31	035.41	0
035	Sodium (%)	0878	0.1650	0.0100	0.1692	0.0138	0.0082	86	-0.31	035.41	0
035	Sodium (%)	0910	0.1650	0.0100	0.1692	0.0138	0.0082	86	-0.31	035.41	0
035	Sodium (%)	0278	0.1650	0.0100	0.1692	0.0138	0.0082	86	-0.31	035.42	0
035	Sodium (%)	0297	0.1650	0.0100	0.1692	0.0138	0.0082	86	-0.31	035.43	0
035	Sodium (%)	0100	0.1650	0.0100	0.1692	0.0138	0.0082	86	-0.31	035.99	0
035	Sodium (%)	0242	0.1650	0.0100	0.1692	0.0138	0.0082	86	-0.31	035.99	0
035	Sodium (%)	0668	0.1660	0.0020	0.1692	0.0138	0.0082	86	-0.24	035.43	0
035	Sodium (%)	2259	0.1680	0.0060	0.1692	0.0138	0.0082	86	-0.09	035.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
035	Sodium (%)	0042	0.1685	0.0050	0.1692	0.0138	0.0082	86	-0.05	035.42	0
035	Sodium (%)	0164	0.1690	0.0040	0.1692	0.0138	0.0082	86	-0.02	035.41	0
035	Sodium (%)	0948	0.1695	0.0050	0.1692	0.0138	0.0082	86	0.02	035.31	0
035	Sodium (%)	0148	0.1700	0.0400	0.1692	0.0138	0.0082	86	0.05	035.41	0
035	Sodium (%)	0693	0.1700	0.0040	0.1692	0.0138	0.0082	86	0.05	035.42	0
035	Sodium (%)	2196	0.1700	0.0000	0.1692	0.0138	0.0082	86	0.05	035.31	0
035	Sodium (%)	0049	0.1700	0.0000	0.1692	0.0138	0.0082	86	0.05	035.41	0
035	Sodium (%)	0186	0.1700	0.0020	0.1692	0.0138	0.0082	86	0.05	035.42	0
035	Sodium (%)	0726	0.1700	0.0000	0.1692	0.0138	0.0082	86	0.05	035.42	0
035	Sodium (%)	0083	0.1700	0.0000	0.1692	0.0138	0.0082	86	0.05	035.43	0
035	Sodium (%)	0675	0.1700	0.0000	0.1692	0.0138	0.0082	86	0.05	035.43	0
035	Sodium (%)	0682	0.1700	0.0000	0.1692	0.0138	0.0082	86	0.05	035.43	0
035	Sodium (%)	0208	0.1710	0.0040	0.1692	0.0138	0.0082	86	0.13	035.41	0
035	Sodium (%)	0229	0.1715	0.0050	0.1692	0.0138	0.0082	86	0.16	035.41	0
035	Sodium (%)	0622	0.1717	0.0046	0.1692	0.0138	0.0082	86	0.18	035.31	0
035	Sodium (%)	0653	0.1740	0.0000	0.1692	0.0138	0.0082	86	0.34	035.41	0
035	Sodium (%)	0098	0.1750	0.0100	0.1692	0.0138	0.0082	86	0.42	035.41	0
035	Sodium (%)	0619	0.1750	0.0040	0.1692	0.0138	0.0082	86	0.42	035.41	0
035	Sodium (%)	0190	0.1750	0.0100	0.1692	0.0138	0.0082	86	0.42	035.42	0
035	Sodium (%)	2192	0.1750	0.0100	0.1692	0.0138	0.0082	86	0.42	035.43	0
035	Sodium (%)	0208	0.1755	0.0010	0.1692	0.0138	0.0082	86	0.45	035.31	0
035	Sodium (%)	0186	0.1755	0.0010	0.1692	0.0138	0.0082	86	0.45	035.52	0
035	Sodium (%)	0520	0.1760	0.0006	0.1692	0.0138	0.0082	86	0.49	035.41	0
035	Sodium (%)	0366	0.1770	0.0120	0.1692	0.0138	0.0082	86	0.56	035.42	0
035	Sodium (%)	0142	0.1771	0.0043	0.1692	0.0138	0.0082	86	0.57	035.31	0
035	Sodium (%)	0202	0.1780	0.0000	0.1692	0.0138	0.0082	86	0.64	035.43	0
035	Sodium (%)	0004	0.1800	0.0200	0.1692	0.0138	0.0082	86	0.78	035.41	0
035	Sodium (%)	0505	0.1800	0.0000	0.1692	0.0138	0.0082	86	0.78	035.43	0
035	Sodium (%)	0651	0.1822	0.0088	0.1692	0.0138	0.0082	86	0.94	035.31	0
035	Sodium (%)	0358	0.1850	0.0100	0.1692	0.0138	0.0082	86	1.14	035.41	0
035	Sodium (%)	0014	0.1880	0.0180	0.1692	0.0138	0.0082	86	1.36	035.31	0
035	Sodium (%)	0226	0.1900	0.0200	0.1692	0.0138	0.0082	86	1.51	035.41	0
035	Sodium (%)	0610	0.1900	0.0000	0.1692	0.0138	0.0082	86	1.51	035.43	0
035	Sodium (%)	0038	0.1905	0.0130	0.1692	0.0138	0.0082	86	1.54	035.43	0
035	Sodium (%)	0263	0.1909	0.0006	0.1692	0.0138	0.0082	86	1.57	035.42	0
035	Sodium (%)	0692	0.1950	0.0300	0.1692	0.0138	0.0082	86	1.87	035.42	0
035	Sodium (%)	0169	0.2050	0.0100	0.1692	0.0138	0.0082	86	2.59	035.32	0
035	Sodium (%)	0689	0.2200	0.0000	0.1692	0.0138	0.0082	86	3.68	035.31	0
035	Sodium (%)	0139	0.2270	0.0000	0.1692	0.0138	0.0082	86	4.19	035.41	0
035	Sodium (%)	0964	0.2296	0.0073	0.1692	0.0138	0.0082	86	4.37	035.43	0
035	Sodium (%)	2146	0.6650	0.0100	0.1692	0.0138	0.0082	86	35.95	035.31	0
035	Sodium (%)	2146	0.7250	0.0100	0.1692	0.0138	0.0082	86	40.30	035.51	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
035	Sodium (%)	0590	1.740	0.0200	0.1692	0.0138	0.0082	86	113.91	035.99	0
035	Sodium (%)	0921	1.770	0.1320	0.1692	0.0138	0.0082	86	116.09	035.31	0
035	Sodium (%)	0510	150.5	1.000	0.1692	0.0138	0.0082	86	10901.88	035.43	2
036	Sulfur (%)	0407	0.2408	0.0001	0.3202	0.0248	0.0087	45	-3.20	036.43	0
036	Sulfur (%)	0870	0.2712	0.0100	0.3202	0.0248	0.0087	45	-1.97	036.42	0
036	Sulfur (%)	0366	0.2750	0.0100	0.3202	0.0248	0.0087	45	-1.82	036.42	0
036	Sulfur (%)	0229	0.2850	0.0100	0.3202	0.0248	0.0087	45	-1.42	036.04	0
036	Sulfur (%)	0242	0.2900	0.0000	0.3202	0.0248	0.0087	45	-1.22	036.99	0
036	Sulfur (%)	0186	0.2935	0.0030	0.3202	0.0248	0.0087	45	-1.07	036.42	0
036	Sulfur (%)	0001	0.2940	0.0180	0.3202	0.0248	0.0087	45	-1.05	036.04	0
036	Sulfur (%)	0226	0.2950	0.0100	0.3202	0.0248	0.0087	45	-1.01	036.04	0
036	Sulfur (%)	0353	0.2950	0.0100	0.3202	0.0248	0.0087	45	-1.01	036.43	0
036	Sulfur (%)	0229	0.3000	0.0000	0.3202	0.0248	0.0087	45	-0.81	036.42	0
036	Sulfur (%)	0032	0.3032	0.0241	0.3202	0.0248	0.0087	45	-0.69	036.43	0
036	Sulfur (%)	0357	0.3037	0.0016	0.3202	0.0248	0.0087	45	-0.66	036.42	0
036	Sulfur (%)	0045	0.3040	0.0100	0.3202	0.0248	0.0087	45	-0.65	036.42	0
036	Sulfur (%)	0035	0.3050	0.0100	0.3202	0.0248	0.0087	45	-0.61	036.43	0
036	Sulfur (%)	0035	0.3100	0.0200	0.3202	0.0248	0.0087	45	-0.41	036.42	0
036	Sulfur (%)	0265	0.3100	0.0000	0.3202	0.0248	0.0087	45	-0.41	036.42	0
036	Sulfur (%)	0297	0.3100	0.0200	0.3202	0.0248	0.0087	45	-0.41	036.43	0
036	Sulfur (%)	0083	0.3150	0.0100	0.3202	0.0248	0.0087	45	-0.21	036.43	0
036	Sulfur (%)	2129	0.3180	0.0101	0.3202	0.0248	0.0087	45	-0.09	036.42	0
036	Sulfur (%)	0345	0.3185	0.0030	0.3202	0.0248	0.0087	45	-0.07	036.43	0
036	Sulfur (%)	2053	0.3200	0.0000	0.3202	0.0248	0.0087	45	-0.01	036.01	0
036	Sulfur (%)	0160	0.3200	0.0000	0.3202	0.0248	0.0087	45	-0.01	036.99	0
036	Sulfur (%)	0693	0.3205	0.0090	0.3202	0.0248	0.0087	45	0.01	036.42	0
036	Sulfur (%)	0186	0.3225	0.0150	0.3202	0.0248	0.0087	45	0.09	036.52	0
036	Sulfur (%)	0202	0.3235	0.0170	0.3202	0.0248	0.0087	45	0.13	036.43	0
036	Sulfur (%)	0726	0.3250	0.0100	0.3202	0.0248	0.0087	45	0.19	036.42	0
036	Sulfur (%)	0038	0.3285	0.0010	0.3202	0.0248	0.0087	45	0.34	036.43	0
036	Sulfur (%)	0148	0.3300	0.0000	0.3202	0.0248	0.0087	45	0.40	036.04	0
036	Sulfur (%)	0610	0.3300	0.0000	0.3202	0.0248	0.0087	45	0.40	036.04	0
036	Sulfur (%)	0164	0.3300	0.0000	0.3202	0.0248	0.0087	45	0.40	036.42	0
036	Sulfur (%)	0692	0.3300	0.0000	0.3202	0.0248	0.0087	45	0.40	036.42	0
036	Sulfur (%)	0037	0.3310	0.0000	0.3202	0.0248	0.0087	45	0.44	036.42	0
036	Sulfur (%)	0668	0.3315	0.0030	0.3202	0.0248	0.0087	45	0.46	036.43	0
036	Sulfur (%)	0708	0.3320	0.0040	0.3202	0.0248	0.0087	45	0.48	036.42	0
036	Sulfur (%)	0560	0.3392	0.0128	0.3202	0.0248	0.0087	45	0.77	036.52	0
036	Sulfur (%)	0278	0.3400	0.0200	0.3202	0.0248	0.0087	45	0.80	036.42	0
036	Sulfur (%)	0027	0.3405	0.0010	0.3202	0.0248	0.0087	45	0.82	036.43	0
036	Sulfur (%)	0098	0.3447	0.0357	0.3202	0.0248	0.0087	45	0.99	036.43	0
036	Sulfur (%)	0169	0.3450	0.0100	0.3202	0.0248	0.0087	45	1.00	036.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
036	Sulfur (%)	0042	0.3470	0.0100	0.3202	0.0248	0.0087	45	1.08	036.42	0
036	Sulfur (%)	0918	0.3510	0.0040	0.3202	0.0248	0.0087	45	1.24	036.43	0
036	Sulfur (%)	0171	0.3550	0.0100	0.3202	0.0248	0.0087	45	1.40	036.42	0
036	Sulfur (%)	0190	0.3650	0.0100	0.3202	0.0248	0.0087	45	1.81	036.42	0
036	Sulfur (%)	0505	0.3650	0.0300	0.3202	0.0248	0.0087	45	1.81	036.43	0
036	Sulfur (%)	0510	0.3850	0.0100	0.3202	0.0248	0.0087	45	2.61	036.43	0
036	Sulfur (%)	0910	3,189	136.0	0.3202	0.0248	0.0087	45	128455.78	036.52	2
037	Zinc (ppm)	0870	155.8	1.700	195.0	13.82	6.003	86	-2.84	037.42	0
037	Zinc (ppm)	0848	158.1	1.800	195.0	13.82	6.003	86	-2.67	037.41	0
037	Zinc (ppm)	0366	168.5	9.000	195.0	13.82	6.003	86	-1.92	037.42	0
037	Zinc (ppm)	0675	174.6	2.330	195.0	13.82	6.003	86	-1.48	037.43	0
037	Zinc (ppm)	0278	175.4	2.600	195.0	13.82	6.003	86	-1.42	037.42	0
037	Zinc (ppm)	0878	176.5	5.000	195.0	13.82	6.003	86	-1.34	037.41	0
037	Zinc (ppm)	0910	177.0	22.00	195.0	13.82	6.003	86	-1.30	037.41	0
037	Zinc (ppm)	0653	177.4	12.90	195.0	13.82	6.003	86	-1.28	037.41	0
037	Zinc (ppm)	0049	177.8	14.10	195.0	13.82	6.003	86	-1.24	037.41	0
037	Zinc (ppm)	0263	177.9	0.0700	195.0	13.82	6.003	86	-1.24	037.42	0
037	Zinc (ppm)	0407	180.3	17.94	195.0	13.82	6.003	86	-1.06	037.41	0
037	Zinc (ppm)	0407	180.4	10.70	195.0	13.82	6.003	86	-1.06	037.43	0
037	Zinc (ppm)	0610	180.5	3.000	195.0	13.82	6.003	86	-1.05	037.43	0
037	Zinc (ppm)	0033	181.0	8.000	195.0	13.82	6.003	86	-1.02	037.43	0
037	Zinc (ppm)	0689	181.8	1.200	195.0	13.82	6.003	86	-0.96	037.31	0
037	Zinc (ppm)	0345	182.0	2.000	195.0	13.82	6.003	86	-0.94	037.43	0
037	Zinc (ppm)	0354	182.6	0.9600	195.0	13.82	6.003	86	-0.90	037.41	0
037	Zinc (ppm)	0148	183.0	2.000	195.0	13.82	6.003	86	-0.87	037.41	0
037	Zinc (ppm)	0045	184.0	8.000	195.0	13.82	6.003	86	-0.80	037.42	0
037	Zinc (ppm)	0918	185.2	3.200	195.0	13.82	6.003	86	-0.71	037.43	0
037	Zinc (ppm)	0074	185.5	3.000	195.0	13.82	6.003	86	-0.69	037.41	0
037	Zinc (ppm)	0265	186.5	7.000	195.0	13.82	6.003	86	-0.62	037.42	0
037	Zinc (ppm)	0668	187.0	0.0000	195.0	13.82	6.003	86	-0.58	037.43	0
037	Zinc (ppm)	0939	187.5	1.000	195.0	13.82	6.003	86	-0.55	037.31	0
037	Zinc (ppm)	0297	187.5	1.000	195.0	13.82	6.003	86	-0.55	037.43	0
037	Zinc (ppm)	0001	188.0	9.900	195.0	13.82	6.003	86	-0.51	037.31	0
037	Zinc (ppm)	0529	188.2	6.700	195.0	13.82	6.003	86	-0.50	037.31	0
037	Zinc (ppm)	2196	188.3	0.0000	195.0	13.82	6.003	86	-0.49	037.31	0
037	Zinc (ppm)	2129	188.3	13.00	195.0	13.82	6.003	86	-0.49	037.42	0
037	Zinc (ppm)	0357	188.7	2.555	195.0	13.82	6.003	86	-0.46	037.42	0
037	Zinc (ppm)	0164	189.5	9.000	195.0	13.82	6.003	86	-0.40	037.44	0
037	Zinc (ppm)	0098	189.9	11.10	195.0	13.82	6.003	86	-0.38	037.41	0
037	Zinc (ppm)	0505	190.4	12.50	195.0	13.82	6.003	86	-0.34	037.43	0
037	Zinc (ppm)	0939	190.5	11.00	195.0	13.82	6.003	86	-0.33	037.53	0
037	Zinc (ppm)	0598	190.8	2.500	195.0	13.82	6.003	86	-0.31	037.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
037	Zinc (ppm)	0563	191.9	2.325	195.0	13.82	6.003	86	-0.23	037.31	0
037	Zinc (ppm)	0511	192.0	2.000	195.0	13.82	6.003	86	-0.22	037.41	0
037	Zinc (ppm)	0208	192.1	0.0000	195.0	13.82	6.003	86	-0.21	037.41	0
037	Zinc (ppm)	0202	192.6	8.500	195.0	13.82	6.003	86	-0.18	037.43	0
037	Zinc (ppm)	0171	193.3	0.2000	195.0	13.82	6.003	86	-0.13	037.41	0
037	Zinc (ppm)	0017	193.5	21.07	195.0	13.82	6.003	86	-0.11	037.43	0
037	Zinc (ppm)	0948	193.5	0.2170	195.0	13.82	6.003	86	-0.11	037.33	0
037	Zinc (ppm)	0242	194.5	13.00	195.0	13.82	6.003	86	-0.04	037.99	0
037	Zinc (ppm)	0010	195.0	10.00	195.0	13.82	6.003	86	0.00	037.33	0
037	Zinc (ppm)	0560	195.0	16.60	195.0	13.82	6.003	86	0.00	037.52	0
037	Zinc (ppm)	0682	195.4	0.0000	195.0	13.82	6.003	86	0.02	037.43	0
037	Zinc (ppm)	2062	196.2	3.600	195.0	13.82	6.003	86	0.08	037.31	0
037	Zinc (ppm)	0047	196.3	0.9300	195.0	13.82	6.003	86	0.09	037.52	0
037	Zinc (ppm)	0358	196.4	11.85	195.0	13.82	6.003	86	0.10	037.41	0
037	Zinc (ppm)	0405	196.8	5.090	195.0	13.82	6.003	86	0.12	037.41	0
037	Zinc (ppm)	0186	198.0	10.00	195.0	13.82	6.003	86	0.21	037.42	0
037	Zinc (ppm)	0693	198.3	3.050	195.0	13.82	6.003	86	0.24	037.42	0
037	Zinc (ppm)	0353	199.0	4.000	195.0	13.82	6.003	86	0.29	037.43	0
037	Zinc (ppm)	0692	199.5	13.00	195.0	13.82	6.003	86	0.32	037.42	0
037	Zinc (ppm)	0160	199.5	11.00	195.0	13.82	6.003	86	0.32	037.44	0
037	Zinc (ppm)	0019	199.8	0.1000	195.0	13.82	6.003	86	0.34	037.41	0
037	Zinc (ppm)	0186	200.0	12.00	195.0	13.82	6.003	86	0.36	037.52	0
037	Zinc (ppm)	2188	200.1	0.0000	195.0	13.82	6.003	86	0.36	037.31	0
037	Zinc (ppm)	0032	200.3	10.10	195.0	13.82	6.003	86	0.38	037.43	0
037	Zinc (ppm)	0726	202.4	5.500	195.0	13.82	6.003	86	0.53	037.42	0
037	Zinc (ppm)	0098	202.4	9.000	195.0	13.82	6.003	86	0.53	037.43	0
037	Zinc (ppm)	0083	202.5	1.000	195.0	13.82	6.003	86	0.54	037.43	0
037	Zinc (ppm)	0226	202.6	13.10	195.0	13.82	6.003	86	0.55	037.41	0
037	Zinc (ppm)	0208	203.5	5.000	195.0	13.82	6.003	86	0.61	037.31	0
037	Zinc (ppm)	0175	205.0	6.000	195.0	13.82	6.003	86	0.72	037.31	0
037	Zinc (ppm)	0590	205.0	4.000	195.0	13.82	6.003	86	0.72	037.99	0
037	Zinc (ppm)	2192	205.5	13.01	195.0	13.82	6.003	86	0.76	037.43	0
037	Zinc (ppm)	0042	206.5	1.000	195.0	13.82	6.003	86	0.83	037.42	0
037	Zinc (ppm)	0622	206.8	4.015	195.0	13.82	6.003	86	0.85	037.31	0
037	Zinc (ppm)	0510	207.0	2.000	195.0	13.82	6.003	86	0.87	037.43	0
037	Zinc (ppm)	2053	207.5	5.000	195.0	13.82	6.003	86	0.90	037.42	0
037	Zinc (ppm)	0038	207.5	3.000	195.0	13.82	6.003	86	0.90	037.43	0
037	Zinc (ppm)	0536	209.1	1.050	195.0	13.82	6.003	86	1.02	037.31	0
037	Zinc (ppm)	0520	209.5	3.000	195.0	13.82	6.003	86	1.05	037.41	0
037	Zinc (ppm)	0229	210.4	5.600	195.0	13.82	6.003	86	1.11	037.41	0
037	Zinc (ppm)	0004	210.8	16.62	195.0	13.82	6.003	86	1.14	037.41	0
037	Zinc (ppm)	0968	212.5	1.000	195.0	13.82	6.003	86	1.26	037.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
037	Zinc (ppm)	0027	215.0	10.00	195.0	13.82	6.003	86	1.44	037.43	0
037	Zinc (ppm)	0190	215.7	1.180	195.0	13.82	6.003	86	1.49	037.42	0
037	Zinc (ppm)	0100	216.0	8.000	195.0	13.82	6.003	86	1.52	037.99	0
037	Zinc (ppm)	0037	216.2	5.500	195.0	13.82	6.003	86	1.53	037.42	0
037	Zinc (ppm)	0169	216.5	1.000	195.0	13.82	6.003	86	1.55	037.43	0
037	Zinc (ppm)	0035	217.0	0.0000	195.0	13.82	6.003	86	1.59	037.42	0
037	Zinc (ppm)	0035	226.0	6.000	195.0	13.82	6.003	86	2.24	037.43	0
037	Zinc (ppm)	0512	230.2	6.300	195.0	13.82	6.003	86	2.54	037.43	0
037	Zinc (ppm)	0619	244.0	2.000	195.0	13.82	6.003	86	3.54	037.41	0
037	Zinc (ppm)	0003	238.0	76.00	195.0	13.82	6.003	86	3.11	037.41	1
038	Molybdenum (ppm)	0297	0.6550	0.0700	3.207	0.5268	0.0997	20	-4.84	038.43	0
038	Molybdenum (ppm)	0353	1.104	0.1596	3.207	0.5268	0.0997	20	-3.99	038.43	0
038	Molybdenum (ppm)	0171	2.510	0.1000	3.207	0.5268	0.0997	20	-1.32	038.41	0
038	Molybdenum (ppm)	0345	2.750	0.0600	3.207	0.5268	0.0997	20	-0.87	038.43	0
038	Molybdenum (ppm)	0407	2.919	0.0234	3.207	0.5268	0.0997	20	-0.55	038.41	0
038	Molybdenum (ppm)	0407	3.036	0.0023	3.207	0.5268	0.0997	20	-0.32	038.43	0
038	Molybdenum (ppm)	0226	3.095	0.0300	3.207	0.5268	0.0997	20	-0.21	038.41	0
038	Molybdenum (ppm)	0169	3.095	0.0100	3.207	0.5268	0.0997	20	-0.21	038.43	0
038	Molybdenum (ppm)	0186	3.120	0.1600	3.207	0.5268	0.0997	20	-0.17	038.52	0
038	Molybdenum (ppm)	0918	3.232	0.0881	3.207	0.5268	0.0997	20	0.05	038.53	0
038	Molybdenum (ppm)	0910	3.235	0.0300	3.207	0.5268	0.0997	20	0.05	038.52	0
038	Molybdenum (ppm)	0083	3.250	0.1000	3.207	0.5268	0.0997	20	0.08	038.43	0
038	Molybdenum (ppm)	0693	3.314	0.2490	3.207	0.5268	0.0997	20	0.20	038.42	0
038	Molybdenum (ppm)	0939	3.355	0.0300	3.207	0.5268	0.0997	20	0.28	038.53	0
038	Molybdenum (ppm)	0038	3.366	0.2120	3.207	0.5268	0.0997	20	0.30	038.43	0
038	Molybdenum (ppm)	0227	3.540	0.1800	3.207	0.5268	0.0997	20	0.63	038.53	0
038	Molybdenum (ppm)	0278	3.675	0.0300	3.207	0.5268	0.0997	20	0.89	038.42	0
038	Molybdenum (ppm)	0045	3.820	0.1600	3.207	0.5268	0.0997	20	1.16	038.42	0
038	Molybdenum (ppm)	0510	4.050	0.1000	3.207	0.5268	0.0997	20	1.60	038.43	0
038	Molybdenum (ppm)	0265	4.600	0.2000	3.207	0.5268	0.0997	20	2.64	038.42	0
040	Barium (ppm)	0918	3.677	0.1287				2		040.53	0
040	Barium (ppm)	0560	3.915	0.1100				2		040.52	0
042	Chloride (%)	0160	0.2100	0.0200				3		042.99	0
042	Chloride (%)	0948	0.2490	0.0120				3		042.00	0
042	Chloride (%)	0297	1.000	0.1400				3		042.99	0
101	Choline Chloride (ppm)	0227	1,855	10.00				1		101.02	0
102	Niacin (ppm)	0227	80.30	1.600				1		102.01	0
103	Pantothenic Acid (ppm)	0227	25.00	0.2000				1		103.01	0
104	Riboflavin (ppm)	0227	19.65	0.9000				3		104.00	0
104	Riboflavin (ppm)	0939	20.01	1.990				3		104.03	0
104	Riboflavin (ppm)	0171	20.05	2.900				3		104.00	0
104	Riboflavin (ppm)	0910	12.99	0.3300				3		104.03	2

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
105	Thiamine (ppm)	0939	7.165	0.0500				3		105.00	0
105	Thiamine (ppm)	0910	7.455	1.330				3		105.00	0
105	Thiamine (ppm)	0227	9.000	0.3800				3		105.01	0
106	Vitamin A (KU / kg)	0619	4.100	1.600	21.89	6.658	2.594	20	-2.67	106.02	0
106	Vitamin A (KU / kg)	0169	4.500	1.000	21.89	6.658	2.594	20	-2.61	106.02	0
106	Vitamin A (KU / kg)	0171	14.85	12.50	21.89	6.658	2.594	20	-1.06	106.00	0
106	Vitamin A (KU / kg)	0905	15.18	0.2540	21.89	6.658	2.594	20	-1.01	106.02	0
106	Vitamin A (KU / kg)	2053	17.60	0.2000	21.89	6.658	2.594	20	-0.64	106.02	0
106	Vitamin A (KU / kg)	0004	19.73	0.2200	21.89	6.658	2.594	20	-0.32	106.02	0
106	Vitamin A (KU / kg)	0610	20.40	0.4000	21.89	6.658	2.594	20	-0.22	106.02	0
106	Vitamin A (KU / kg)	0676	20.60	0.2000	21.89	6.658	2.594	20	-0.19	106.02	0
106	Vitamin A (KU / kg)	0689	20.85	0.9000	21.89	6.658	2.594	20	-0.16	106.02	0
106	Vitamin A (KU / kg)	2103	21.22	0.5200	21.89	6.658	2.594	20	-0.10	106.02	0
106	Vitamin A (KU / kg)	0638	21.70	0.0000	21.89	6.658	2.594	20	-0.03	106.02	0
106	Vitamin A (KU / kg)	0227	21.85	2.300	21.89	6.658	2.594	20	-0.01	106.02	0
106	Vitamin A (KU / kg)	0910	22.55	1.700	21.89	6.658	2.594	20	0.10	106.02	0
106	Vitamin A (KU / kg)	0010	23.50	3.700	21.89	6.658	2.594	20	0.24	106.02	0
106	Vitamin A (KU / kg)	0675	25.66	5.490	21.89	6.658	2.594	20	0.57	106.02	0
106	Vitamin A (KU / kg)	0941	27.66	8.550	21.89	6.658	2.594	20	0.87	106.02	0
106	Vitamin A (KU / kg)	0563	28.34	0.0233	21.89	6.658	2.594	20	0.97	106.02	0
106	Vitamin A (KU / kg)	0208	29.54	4.409	21.89	6.658	2.594	20	1.15	106.02	0
106	Vitamin A (KU / kg)	0098	30.95	0.9000	21.89	6.658	2.594	20	1.36	106.01	0
106	Vitamin A (KU / kg)	2066	31.92	7.010	21.89	6.658	2.594	20	1.51	106.02	0
106	Vitamin A (KU / kg)	0148	22,213	456.2	21.89	6.658	2.594	20	3332.97	106.02	2
107	Vitamin B12 (ppb)	0227	59.10	5.800				1		107.00	0
108	Vitamin D3 (KU / kg)	0169	4.000	0.0000	15.01	3.488	0.6498	11	-3.16	108.01	0
108	Vitamin D3 (KU / kg)	0675	11.25	0.2900	15.01	3.488	0.6498	11	-1.08	108.02	0
108	Vitamin D3 (KU / kg)	0610	12.30	1.400	15.01	3.488	0.6498	11	-0.78	108.02	0
108	Vitamin D3 (KU / kg)	0171	12.45	1.500	15.01	3.488	0.6498	11	-0.73	108.99	0
108	Vitamin D3 (KU / kg)	0619	15.05	0.3000	15.01	3.488	0.6498	11	0.01	108.99	0
108	Vitamin D3 (KU / kg)	2192	16.14	1.758	15.01	3.488	0.6498	11	0.33	108.02	0
108	Vitamin D3 (KU / kg)	2103	16.53	0.5800	15.01	3.488	0.6498	11	0.44	108.02	0
108	Vitamin D3 (KU / kg)	0010	16.62	0.2200	15.01	3.488	0.6498	11	0.46	108.99	0
108	Vitamin D3 (KU / kg)	0910	18.15	0.7000	15.01	3.488	0.6498	11	0.90	108.02	0
108	Vitamin D3 (KU / kg)	0227	18.20	0.4000	15.01	3.488	0.6498	11	0.92	108.02	0
108	Vitamin D3 (KU / kg)	0676	18.60	0.0000	15.01	3.488	0.6498	11	1.03	108.02	0
108	Vitamin D3 (KU / kg)	0208	16.90	4.000	15.01	3.488	0.6498	11	0.54	108.02	1
109	Vitamin E (IU / kg)	0169	25.50	1.000	171.4	49.50	7.424	16	-2.95	109.02	0
109	Vitamin E (IU / kg)	0675	45.20	8.680	171.4	49.50	7.424	16	-2.55	109.02	0
109	Vitamin E (IU / kg)	0638	99.09	0.0000	171.4	49.50	7.424	16	-1.46	109.02	0
109	Vitamin E (IU / kg)	0905	113.5	19.00	171.4	49.50	7.424	16	-1.17	109.02	0
109	Vitamin E (IU / kg)	0208	176.3	4.556	171.4	49.50	7.424	16	0.10	109.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
109	Vitamin E (IU / kg)	0563	177.0	1.834	171.4	49.50	7.424	16	0.11	109.02	0
109	Vitamin E (IU / kg)	0910	180.5	1.000	171.4	49.50	7.424	16	0.18	109.02	0
109	Vitamin E (IU / kg)	2103	184.0	0.0000	171.4	49.50	7.424	16	0.25	109.02	0
109	Vitamin E (IU / kg)	0941	189.6	17.67	171.4	49.50	7.424	16	0.37	109.02	0
109	Vitamin E (IU / kg)	0098	190.0	0.0000	171.4	49.50	7.424	16	0.38	109.02	0
109	Vitamin E (IU / kg)	0610	193.5	13.00	171.4	49.50	7.424	16	0.45	109.02	0
109	Vitamin E (IU / kg)	0171	194.0	2.000	171.4	49.50	7.424	16	0.46	109.99	0
109	Vitamin E (IU / kg)	0148	202.5	38.04	171.4	49.50	7.424	16	0.63	109.02	0
109	Vitamin E (IU / kg)	0590	209.5	1.000	171.4	49.50	7.424	16	0.77	109.02	0
109	Vitamin E (IU / kg)	0676	218.5	11.00	171.4	49.50	7.424	16	0.95	109.02	0
109	Vitamin E (IU / kg)	0227	220.0	0.0000	171.4	49.50	7.424	16	0.98	109.02	0
109	Vitamin E (IU / kg)	0619	1,315	110.0	171.4	49.50	7.424	16	23.10	109.02	2
110	Xanthophyll (ppm)	0171	9.450	0.9000				2		110.00	0
110	Xanthophyll (ppm)	0939	16.50	0.5100				2		110.00	0
112	Pyridoxine (µg / g)	0227	9.185	0.3300				1		112.01	0
113	Folic Acid (ppm)	0227	7.275	0.4100				1		113.01	0
114	Biotin (ppm)	0227	0.6560	0.0000				1		114.01	0
115	Non Protein N (NPN) (%)	0098	0.5050	0.0700				1		115.00	0
118	Peroxide value (meq/kg)	2081	1.800	0.2800				1		118.99	0
120	Alanine (%)	0889	0.7750	0.0000	0.9560	0.0557	0.0133	29	-3.25	120.99	0
120	Alanine (%)	0148	0.8500	0.0000	0.9560	0.0557	0.0133	29	-1.90	120.05	0
120	Alanine (%)	0626	0.8520	0.0160	0.9560	0.0557	0.0133	29	-1.87	120.05	0
120	Alanine (%)	0675	0.8650	0.0100	0.9560	0.0557	0.0133	29	-1.64	120.00	0
120	Alanine (%)	0353	0.8900	0.0200	0.9560	0.0557	0.0133	29	-1.19	120.00	0
120	Alanine (%)	0910	0.8900	0.0600	0.9560	0.0557	0.0133	29	-1.19	120.00	0
120	Alanine (%)	0610	0.8950	0.0100	0.9560	0.0557	0.0133	29	-1.10	120.05	0
120	Alanine (%)	0918	0.9352	0.0022	0.9560	0.0557	0.0133	29	-0.37	120.00	0
120	Alanine (%)	0208	0.9450	0.0100	0.9560	0.0557	0.0133	29	-0.20	120.00	0
120	Alanine (%)	0968	0.9600	0.0000	0.9560	0.0557	0.0133	29	0.07	120.00	0
120	Alanine (%)	0676	0.9600	0.0200	0.9560	0.0557	0.0133	29	0.07	120.05	0
120	Alanine (%)	0870	0.9604	0.0032	0.9560	0.0557	0.0133	29	0.08	120.00	0
120	Alanine (%)	0571	0.9610	0.0160	0.9560	0.0557	0.0133	29	0.09	120.00	0
120	Alanine (%)	0878	0.9630	0.0080	0.9560	0.0557	0.0133	29	0.12	120.00	0
120	Alanine (%)	2196	0.9640	0.0000	0.9560	0.0557	0.0133	29	0.14	120.05	0
120	Alanine (%)	0652	0.9650	0.0300	0.9560	0.0557	0.0133	29	0.16	120.00	0
120	Alanine (%)	0098	0.9685	0.0190	0.9560	0.0557	0.0133	29	0.22	120.02	0
120	Alanine (%)	0644	0.9700	0.0020	0.9560	0.0557	0.0133	29	0.25	120.00	0
120	Alanine (%)	2059	0.9755	0.0050	0.9560	0.0557	0.0133	29	0.35	120.00	0
120	Alanine (%)	0872	0.9790	0.0000	0.9560	0.0557	0.0133	29	0.41	120.00	0
120	Alanine (%)	0684	0.9845	0.0210	0.9560	0.0557	0.0133	29	0.51	120.00	0
120	Alanine (%)	0682	0.9910	0.0000	0.9560	0.0557	0.0133	29	0.63	120.00	0
120	Alanine (%)	0227	0.9950	0.0100	0.9560	0.0557	0.0133	29	0.70	120.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
120	Alanine (%)	0242	0.9950	0.0100	0.9560	0.0557	0.0133	29	0.70	120.05	0
120	Alanine (%)	0619	1.000	0.0000	0.9560	0.0557	0.0133	29	0.79	120.00	0
120	Alanine (%)	0859	1.004	0.0019	0.9560	0.0557	0.0133	29	0.86	120.00	0
120	Alanine (%)	0171	1.019	0.0130	0.9560	0.0557	0.0133	29	1.12	120.00	0
120	Alanine (%)	0407	1.026	0.0340	0.9560	0.0557	0.0133	29	1.26	120.05	0
120	Alanine (%)	2188	1.058	0.0650	0.9560	0.0557	0.0133	29	1.82	120.05	0
120	Alanine (%)	0226	0.9221	0.0957	0.9560	0.0557	0.0133	29	-0.61	120.00	1
120	Alanine (%)	0354	1.602	0.0380	0.9560	0.0557	0.0133	29	11.60	120.00	2
121	Arginine (%)	0889	0.4650	0.0000	1.073	0.0582	0.0173	30	-10.43	121.99	0
121	Arginine (%)	0353	0.9300	0.0000	1.073	0.0582	0.0173	30	-2.45	121.00	0
121	Arginine (%)	0626	0.9905	0.0130	1.073	0.0582	0.0173	30	-1.41	121.05	0
121	Arginine (%)	0675	1.005	0.0100	1.073	0.0582	0.0173	30	-1.16	121.00	0
121	Arginine (%)	0870	1.012	0.0069	1.073	0.0582	0.0173	30	-1.04	121.00	0
121	Arginine (%)	0968	1.025	0.0100	1.073	0.0582	0.0173	30	-0.82	121.00	0
121	Arginine (%)	0208	1.030	0.0000	1.073	0.0582	0.0173	30	-0.73	121.00	0
121	Arginine (%)	0676	1.030	0.0200	1.073	0.0582	0.0173	30	-0.73	121.05	0
121	Arginine (%)	0918	1.037	0.0227	1.073	0.0582	0.0173	30	-0.61	121.00	0
121	Arginine (%)	0227	1.040	0.0200	1.073	0.0582	0.0173	30	-0.56	121.00	0
121	Arginine (%)	0910	1.045	0.0100	1.073	0.0582	0.0173	30	-0.48	121.00	0
121	Arginine (%)	0610	1.050	0.0000	1.073	0.0582	0.0173	30	-0.39	121.05	0
121	Arginine (%)	0098	1.065	0.0440	1.073	0.0582	0.0173	30	-0.13	121.02	0
121	Arginine (%)	0652	1.070	0.0200	1.073	0.0582	0.0173	30	-0.05	121.00	0
121	Arginine (%)	0354	1.072	0.0200	1.073	0.0582	0.0173	30	-0.01	121.00	0
121	Arginine (%)	0571	1.077	0.0060	1.073	0.0582	0.0173	30	0.07	121.00	0
121	Arginine (%)	0872	1.091	0.0070	1.073	0.0582	0.0173	30	0.31	121.00	0
121	Arginine (%)	0859	1.094	0.0169	1.073	0.0582	0.0173	30	0.37	121.00	0
121	Arginine (%)	2059	1.096	0.0020	1.073	0.0582	0.0173	30	0.40	121.00	0
121	Arginine (%)	0878	1.099	0.0330	1.073	0.0582	0.0173	30	0.44	121.00	0
121	Arginine (%)	0242	1.100	0.0200	1.073	0.0582	0.0173	30	0.47	121.05	0
121	Arginine (%)	2188	1.108	0.0210	1.073	0.0582	0.0173	30	0.60	121.05	0
121	Arginine (%)	0644	1.113	0.0020	1.073	0.0582	0.0173	30	0.69	121.00	0
121	Arginine (%)	2196	1.119	0.0000	1.073	0.0582	0.0173	30	0.79	121.05	0
121	Arginine (%)	0684	1.124	0.0070	1.073	0.0582	0.0173	30	0.87	121.00	0
121	Arginine (%)	0148	1.130	0.1000	1.073	0.0582	0.0173	30	0.98	121.05	0
121	Arginine (%)	0171	1.132	0.0400	1.073	0.0582	0.0173	30	1.02	121.00	0
121	Arginine (%)	0682	1.138	0.0000	1.073	0.0582	0.0173	30	1.12	121.00	0
121	Arginine (%)	0619	1.160	0.0000	1.073	0.0582	0.0173	30	1.50	121.00	0
121	Arginine (%)	0407	1.987	0.0680	1.073	0.0582	0.0173	30	15.70	121.05	0
121	Arginine (%)	0226	1.139	0.1102	1.073	0.0582	0.0173	30	1.13	121.00	1
122	Aspartic (%)	0148	1.445	0.0300	1.606	0.0778	0.0194	30	-2.07	122.05	0
122	Aspartic (%)	0675	1.465	0.0100	1.606	0.0778	0.0194	30	-1.81	122.00	0
122	Aspartic (%)	0407	1.487	0.0290	1.606	0.0778	0.0194	30	-1.54	122.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
122	Aspartic (%)	0889	1.515	0.0300	1.606	0.0778	0.0194	30	-1.17	122.99	0
122	Aspartic (%)	0354	1.524	0.0010	1.606	0.0778	0.0194	30	-1.06	122.00	0
122	Aspartic (%)	0918	1.535	0.0363	1.606	0.0778	0.0194	30	-0.91	122.00	0
122	Aspartic (%)	0910	1.545	0.0300	1.606	0.0778	0.0194	30	-0.78	122.00	0
122	Aspartic (%)	0208	1.555	0.0100	1.606	0.0778	0.0194	30	-0.66	122.00	0
122	Aspartic (%)	0610	1.560	0.0400	1.606	0.0778	0.0194	30	-0.59	122.05	0
122	Aspartic (%)	0870	1.574	0.0398	1.606	0.0778	0.0194	30	-0.41	122.00	0
122	Aspartic (%)	0571	1.580	0.0070	1.606	0.0778	0.0194	30	-0.34	122.00	0
122	Aspartic (%)	0353	1.600	0.0200	1.606	0.0778	0.0194	30	-0.08	122.00	0
122	Aspartic (%)	0878	1.606	0.0010	1.606	0.0778	0.0194	30	-0.01	122.00	0
122	Aspartic (%)	0626	1.608	0.0240	1.606	0.0778	0.0194	30	0.03	122.05	0
122	Aspartic (%)	2059	1.616	0.0160	1.606	0.0778	0.0194	30	0.13	122.00	0
122	Aspartic (%)	0644	1.621	0.0040	1.606	0.0778	0.0194	30	0.19	122.00	0
122	Aspartic (%)	0652	1.625	0.0100	1.606	0.0778	0.0194	30	0.24	122.00	0
122	Aspartic (%)	0684	1.628	0.0190	1.606	0.0778	0.0194	30	0.28	122.00	0
122	Aspartic (%)	0872	1.629	0.0020	1.606	0.0778	0.0194	30	0.30	122.00	0
122	Aspartic (%)	0227	1.630	0.0000	1.606	0.0778	0.0194	30	0.31	122.00	0
122	Aspartic (%)	0098	1.637	0.0680	1.606	0.0778	0.0194	30	0.40	122.02	0
122	Aspartic (%)	0859	1.638	0.0166	1.606	0.0778	0.0194	30	0.41	122.00	0
122	Aspartic (%)	0682	1.641	0.0000	1.606	0.0778	0.0194	30	0.45	122.00	0
122	Aspartic (%)	2196	1.641	0.0000	1.606	0.0778	0.0194	30	0.45	122.05	0
122	Aspartic (%)	0968	1.645	0.0100	1.606	0.0778	0.0194	30	0.50	122.00	0
122	Aspartic (%)	2188	1.694	0.0010	1.606	0.0778	0.0194	30	1.12	122.05	0
122	Aspartic (%)	0242	1.700	0.0200	1.606	0.0778	0.0194	30	1.21	122.05	0
122	Aspartic (%)	0676	1.725	0.0500	1.606	0.0778	0.0194	30	1.53	122.05	0
122	Aspartic (%)	0171	1.726	0.0460	1.606	0.0778	0.0194	30	1.54	122.00	0
122	Aspartic (%)	0619	1.755	0.0100	1.606	0.0778	0.0194	30	1.91	122.00	0
122	Aspartic (%)	0226	1.709	0.1477	1.606	0.0778	0.0194	30	1.33	122.00	1
124	Cysteine/Cystine (%)	0870	0.2649	0.0094	0.3116	0.0302	0.0097	29	-1.55	124.00	0
124	Cysteine/Cystine (%)	0353	0.2700	0.0000	0.3116	0.0302	0.0097	29	-1.38	124.00	0
124	Cysteine/Cystine (%)	0227	0.2750	0.0100	0.3116	0.0302	0.0097	29	-1.21	124.00	0
124	Cysteine/Cystine (%)	0910	0.2800	0.0200	0.3116	0.0302	0.0097	29	-1.05	124.00	0
124	Cysteine/Cystine (%)	0859	0.2875	0.0021	0.3116	0.0302	0.0097	29	-0.80	124.00	0
124	Cysteine/Cystine (%)	0644	0.2875	0.0030	0.3116	0.0302	0.0097	29	-0.80	124.00	0
124	Cysteine/Cystine (%)	0652	0.2950	0.0100	0.3116	0.0302	0.0097	29	-0.55	124.00	0
124	Cysteine/Cystine (%)	2196	0.3000	0.0000	0.3116	0.0302	0.0097	29	-0.38	124.05	0
124	Cysteine/Cystine (%)	0226	0.3006	0.0040	0.3116	0.0302	0.0097	29	-0.36	124.00	0
124	Cysteine/Cystine (%)	0968	0.3010	0.0020	0.3116	0.0302	0.0097	29	-0.35	124.00	0
124	Cysteine/Cystine (%)	0878	0.3015	0.0030	0.3116	0.0302	0.0097	29	-0.33	124.00	0
124	Cysteine/Cystine (%)	2059	0.3015	0.0030	0.3116	0.0302	0.0097	29	-0.33	124.00	0
124	Cysteine/Cystine (%)	0571	0.3020	0.0020	0.3116	0.0302	0.0097	29	-0.32	124.00	0
124	Cysteine/Cystine (%)	0872	0.3020	0.0020	0.3116	0.0302	0.0097	29	-0.32	124.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
124	Cysteine/Cystine (%)	0684	0.3035	0.0130	0.3116	0.0302	0.0097	29	-0.27	124.00	0
124	Cysteine/Cystine (%)	2188	0.3090	0.0040	0.3116	0.0302	0.0097	29	-0.08	124.05	0
124	Cysteine/Cystine (%)	0676	0.3100	0.0000	0.3116	0.0302	0.0097	29	-0.05	124.05	0
124	Cysteine/Cystine (%)	0407	0.3125	0.0022	0.3116	0.0302	0.0097	29	0.03	124.05	0
124	Cysteine/Cystine (%)	0098	0.3130	0.0040	0.3116	0.0302	0.0097	29	0.05	124.02	0
124	Cysteine/Cystine (%)	0208	0.3140	0.0180	0.3116	0.0302	0.0097	29	0.08	124.00	0
124	Cysteine/Cystine (%)	0354	0.3225	0.0050	0.3116	0.0302	0.0097	29	0.36	124.00	0
124	Cysteine/Cystine (%)	0619	0.3285	0.0290	0.3116	0.0302	0.0097	29	0.56	124.00	0
124	Cysteine/Cystine (%)	0610	0.3350	0.0100	0.3116	0.0302	0.0097	29	0.78	124.05	0
124	Cysteine/Cystine (%)	0918	0.3400	0.0200	0.3116	0.0302	0.0097	29	0.94	124.00	0
124	Cysteine/Cystine (%)	0242	0.3500	0.0000	0.3116	0.0302	0.0097	29	1.28	124.05	0
124	Cysteine/Cystine (%)	0148	0.3650	0.0100	0.3116	0.0302	0.0097	29	1.77	124.05	0
124	Cysteine/Cystine (%)	0171	0.3725	0.0410	0.3116	0.0302	0.0097	29	2.02	124.00	0
124	Cysteine/Cystine (%)	0682	0.3820	0.0000	0.3116	0.0302	0.0097	29	2.34	124.00	0
124	Cysteine/Cystine (%)	0889	0.6325	0.0550	0.3116	0.0302	0.0097	29	10.64	124.99	0
124	Cysteine/Cystine (%)	0675	0.4600	0.0800	0.3116	0.0302	0.0097	29	4.92	124.00	1
125	Glutamic (%)	0889	2.475	0.1000	3.210	0.1369	0.0537	31	-5.37	125.99	0
125	Glutamic (%)	0675	2.820	0.1600	3.210	0.1369	0.0537	31	-2.85	125.00	0
125	Glutamic (%)	0171	2.978	0.0560	3.210	0.1369	0.0537	31	-1.70	125.00	0
125	Glutamic (%)	0918	2.987	0.0091	3.210	0.1369	0.0537	31	-1.63	125.00	0
125	Glutamic (%)	0910	3.055	0.0500	3.210	0.1369	0.0537	31	-1.13	125.00	0
125	Glutamic (%)	0610	3.055	0.0500	3.210	0.1369	0.0537	31	-1.13	125.05	0
125	Glutamic (%)	2196	3.137	0.0000	3.210	0.1369	0.0537	31	-0.53	125.05	0
125	Glutamic (%)	2059	3.147	0.0200	3.210	0.1369	0.0537	31	-0.46	125.00	0
125	Glutamic (%)	0571	3.148	0.0450	3.210	0.1369	0.0537	31	-0.46	125.00	0
125	Glutamic (%)	0148	3.155	0.1100	3.210	0.1369	0.0537	31	-0.40	125.05	0
125	Glutamic (%)	0354	3.167	0.0080	3.210	0.1369	0.0537	31	-0.31	125.00	0
125	Glutamic (%)	0626	3.191	0.0530	3.210	0.1369	0.0537	31	-0.14	125.05	0
125	Glutamic (%)	0676	3.200	0.0200	3.210	0.1369	0.0537	31	-0.07	125.05	0
125	Glutamic (%)	0872	3.205	0.0210	3.210	0.1369	0.0537	31	-0.04	125.00	0
125	Glutamic (%)	2188	3.207	0.0360	3.210	0.1369	0.0537	31	-0.02	125.05	0
125	Glutamic (%)	0870	3.220	0.0264	3.210	0.1369	0.0537	31	0.07	125.00	0
125	Glutamic (%)	0652	3.225	0.0700	3.210	0.1369	0.0537	31	0.11	125.00	0
125	Glutamic (%)	0227	3.235	0.0300	3.210	0.1369	0.0537	31	0.18	125.00	0
125	Glutamic (%)	0644	3.241	0.0100	3.210	0.1369	0.0537	31	0.23	125.00	0
125	Glutamic (%)	0968	3.245	0.0100	3.210	0.1369	0.0537	31	0.26	125.00	0
125	Glutamic (%)	0353	3.260	0.0800	3.210	0.1369	0.0537	31	0.36	125.00	0
125	Glutamic (%)	0098	3.276	0.1040	3.210	0.1369	0.0537	31	0.48	125.02	0
125	Glutamic (%)	0682	3.286	0.0000	3.210	0.1369	0.0537	31	0.55	125.00	0
125	Glutamic (%)	0859	3.287	0.0317	3.210	0.1369	0.0537	31	0.56	125.00	0
125	Glutamic (%)	0684	3.293	0.0970	3.210	0.1369	0.0537	31	0.60	125.00	0
125	Glutamic (%)	0878	3.337	0.0050	3.210	0.1369	0.0537	31	0.92	125.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
125	Glutamic (%)	0407	3.344	0.1990	3.210	0.1369	0.0537	31	0.97	125.05	0
125	Glutamic (%)	0242	3.355	0.0100	3.210	0.1369	0.0537	31	1.06	125.05	0
125	Glutamic (%)	0208	3.395	0.0100	3.210	0.1369	0.0537	31	1.35	125.00	0
125	Glutamic (%)	0226	3.481	0.2344	3.210	0.1369	0.0537	31	1.98	125.00	0
125	Glutamic (%)	0619	3.655	0.0100	3.210	0.1369	0.0537	31	3.25	125.00	0
126	Glycine (%)	0407	0.3865	0.0290	0.7465	0.0326	0.0145	30	-11.04	126.05	0
126	Glycine (%)	0675	0.6950	0.0100	0.7465	0.0326	0.0145	30	-1.58	126.00	0
126	Glycine (%)	0910	0.7050	0.0100	0.7465	0.0326	0.0145	30	-1.27	126.00	0
126	Glycine (%)	0918	0.7094	0.0165	0.7465	0.0326	0.0145	30	-1.14	126.00	0
126	Glycine (%)	0353	0.7100	0.0000	0.7465	0.0326	0.0145	30	-1.12	126.00	0
126	Glycine (%)	0208	0.7200	0.0000	0.7465	0.0326	0.0145	30	-0.81	126.00	0
126	Glycine (%)	0610	0.7200	0.0200	0.7465	0.0326	0.0145	30	-0.81	126.05	0
126	Glycine (%)	0626	0.7200	0.0140	0.7465	0.0326	0.0145	30	-0.81	126.05	0
126	Glycine (%)	0870	0.7205	0.0228	0.7465	0.0326	0.0145	30	-0.80	126.00	0
126	Glycine (%)	0571	0.7370	0.0040	0.7465	0.0326	0.0145	30	-0.29	126.00	0
126	Glycine (%)	0148	0.7400	0.0200	0.7465	0.0326	0.0145	30	-0.20	126.05	0
126	Glycine (%)	0684	0.7440	0.0060	0.7465	0.0326	0.0145	30	-0.08	126.00	0
126	Glycine (%)	0354	0.7445	0.0130	0.7465	0.0326	0.0145	30	-0.06	126.00	0
126	Glycine (%)	0652	0.7450	0.0100	0.7465	0.0326	0.0145	30	-0.05	126.00	0
126	Glycine (%)	0968	0.7450	0.0100	0.7465	0.0326	0.0145	30	-0.05	126.00	0
126	Glycine (%)	2059	0.7525	0.0010	0.7465	0.0326	0.0145	30	0.18	126.00	0
126	Glycine (%)	0227	0.7550	0.0100	0.7465	0.0326	0.0145	30	0.26	126.00	0
126	Glycine (%)	0644	0.7555	0.0170	0.7465	0.0326	0.0145	30	0.28	126.00	0
126	Glycine (%)	0682	0.7560	0.0000	0.7465	0.0326	0.0145	30	0.29	126.00	0
126	Glycine (%)	2196	0.7560	0.0000	0.7465	0.0326	0.0145	30	0.29	126.05	0
126	Glycine (%)	0878	0.7590	0.0200	0.7465	0.0326	0.0145	30	0.38	126.00	0
126	Glycine (%)	0098	0.7590	0.0320	0.7465	0.0326	0.0145	30	0.38	126.02	0
126	Glycine (%)	0859	0.7615	0.0035	0.7465	0.0326	0.0145	30	0.46	126.00	0
126	Glycine (%)	0872	0.7645	0.0010	0.7465	0.0326	0.0145	30	0.55	126.00	0
126	Glycine (%)	2188	0.7685	0.0090	0.7465	0.0326	0.0145	30	0.68	126.05	0
126	Glycine (%)	0242	0.7700	0.0000	0.7465	0.0326	0.0145	30	0.72	126.05	0
126	Glycine (%)	0676	0.8000	0.0600	0.7465	0.0326	0.0145	30	1.64	126.05	0
126	Glycine (%)	0171	0.8035	0.0490	0.7465	0.0326	0.0145	30	1.75	126.00	0
126	Glycine (%)	0619	0.8100	0.0020	0.7465	0.0326	0.0145	30	1.95	126.00	0
126	Glycine (%)	0889	1.038	0.0450	0.7465	0.0326	0.0145	30	8.92	126.99	0
126	Glycine (%)	0226	0.7361	0.0923	0.7465	0.0326	0.0145	30	-0.32	126.00	1
127	Histidine (%)	0870	0.4192	0.0153	0.4683	0.0261	0.0117	29	-1.89	127.00	0
127	Histidine (%)	0676	0.4250	0.0300	0.4683	0.0261	0.0117	29	-1.66	127.05	0
127	Histidine (%)	0626	0.4295	0.0010	0.4683	0.0261	0.0117	29	-1.49	127.05	0
127	Histidine (%)	0918	0.4383	0.0083	0.4683	0.0261	0.0117	29	-1.15	127.00	0
127	Histidine (%)	0148	0.4400	0.0400	0.4683	0.0261	0.0117	29	-1.09	127.05	0
127	Histidine (%)	0353	0.4500	0.0200	0.4683	0.0261	0.0117	29	-0.70	127.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
127	Histidine (%)	0910	0.4500	0.0200	0.4683	0.0261	0.0117	29	-0.70	127.00	0
127	Histidine (%)	0968	0.4500	0.0200	0.4683	0.0261	0.0117	29	-0.70	127.00	0
127	Histidine (%)	2188	0.4580	0.0020	0.4683	0.0261	0.0117	29	-0.40	127.05	0
127	Histidine (%)	0208	0.4600	0.0000	0.4683	0.0261	0.0117	29	-0.32	127.00	0
127	Histidine (%)	0571	0.4615	0.0050	0.4683	0.0261	0.0117	29	-0.26	127.00	0
127	Histidine (%)	0098	0.4620	0.0140	0.4683	0.0261	0.0117	29	-0.24	127.02	0
127	Histidine (%)	0610	0.4650	0.0100	0.4683	0.0261	0.0117	29	-0.13	127.05	0
127	Histidine (%)	0859	0.4679	0.0036	0.4683	0.0261	0.0117	29	-0.02	127.00	0
127	Histidine (%)	0878	0.4690	0.0020	0.4683	0.0261	0.0117	29	0.03	127.00	0
127	Histidine (%)	0226	0.4710	0.0249	0.4683	0.0261	0.0117	29	0.10	127.00	0
127	Histidine (%)	0227	0.4750	0.0100	0.4683	0.0261	0.0117	29	0.26	127.00	0
127	Histidine (%)	2059	0.4750	0.0080	0.4683	0.0261	0.0117	29	0.26	127.00	0
127	Histidine (%)	0354	0.4755	0.0010	0.4683	0.0261	0.0117	29	0.28	127.00	0
127	Histidine (%)	0872	0.4785	0.0010	0.4683	0.0261	0.0117	29	0.39	127.00	0
127	Histidine (%)	0684	0.4800	0.0220	0.4683	0.0261	0.0117	29	0.45	127.00	0
127	Histidine (%)	0644	0.4820	0.0000	0.4683	0.0261	0.0117	29	0.53	127.00	0
127	Histidine (%)	2196	0.4830	0.0000	0.4683	0.0261	0.0117	29	0.56	127.05	0
127	Histidine (%)	0407	0.4865	0.0170	0.4683	0.0261	0.0117	29	0.70	127.05	0
127	Histidine (%)	0652	0.4950	0.0100	0.4683	0.0261	0.0117	29	1.02	127.00	0
127	Histidine (%)	0682	0.5000	0.0000	0.4683	0.0261	0.0117	29	1.22	127.00	0
127	Histidine (%)	0242	0.5050	0.0100	0.4683	0.0261	0.0117	29	1.41	127.05	0
127	Histidine (%)	0675	0.5150	0.0100	0.4683	0.0261	0.0117	29	1.79	127.00	0
127	Histidine (%)	0619	0.5655	0.0330	0.4683	0.0261	0.0117	29	3.73	127.00	0
127	Histidine (%)	0171	0.5675	0.0990	0.4683	0.0261	0.0117	29	3.81	127.00	1
127	Histidine (%)	0889	1.038	0.0450	0.4683	0.0261	0.0117	29	21.85	127.99	2
128	Isoleucine (%)	0353	0.5550	0.0100	0.6890	0.0446	0.0173	31	-3.00	128.00	0
128	Isoleucine (%)	0675	0.5750	0.0500	0.6890	0.0446	0.0173	31	-2.55	128.00	0
128	Isoleucine (%)	0226	0.6227	0.0622	0.6890	0.0446	0.0173	31	-1.49	128.00	0
128	Isoleucine (%)	0870	0.6266	0.0133	0.6890	0.0446	0.0173	31	-1.40	128.00	0
128	Isoleucine (%)	0918	0.6271	0.0132	0.6890	0.0446	0.0173	31	-1.39	128.00	0
128	Isoleucine (%)	0619	0.6400	0.0600	0.6890	0.0446	0.0173	31	-1.10	128.00	0
128	Isoleucine (%)	0652	0.6650	0.0100	0.6890	0.0446	0.0173	31	-0.54	128.00	0
128	Isoleucine (%)	0171	0.6720	0.0500	0.6890	0.0446	0.0173	31	-0.38	128.00	0
128	Isoleucine (%)	0968	0.6750	0.0100	0.6890	0.0446	0.0173	31	-0.31	128.00	0
128	Isoleucine (%)	0910	0.6800	0.0400	0.6890	0.0446	0.0173	31	-0.20	128.00	0
128	Isoleucine (%)	0208	0.6800	0.0000	0.6890	0.0446	0.0173	31	-0.20	128.00	0
128	Isoleucine (%)	0626	0.6805	0.0090	0.6890	0.0446	0.0173	31	-0.19	128.05	0
128	Isoleucine (%)	0859	0.6845	0.0089	0.6890	0.0446	0.0173	31	-0.10	128.00	0
128	Isoleucine (%)	0354	0.6875	0.0050	0.6890	0.0446	0.0173	31	-0.03	128.00	0
128	Isoleucine (%)	0878	0.6895	0.0130	0.6890	0.0446	0.0173	31	0.01	128.00	0
128	Isoleucine (%)	0610	0.6900	0.0200	0.6890	0.0446	0.0173	31	0.02	128.05	0
128	Isoleucine (%)	0676	0.6900	0.0200	0.6890	0.0446	0.0173	31	0.02	128.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
128	Isoleucine (%)	2196	0.6950	0.0000	0.6890	0.0446	0.0173	31	0.13	128.05	0
128	Isoleucine (%)	0098	0.6995	0.0190	0.6890	0.0446	0.0173	31	0.23	128.02	0
128	Isoleucine (%)	0682	0.7020	0.0000	0.6890	0.0446	0.0173	31	0.29	128.00	0
128	Isoleucine (%)	0684	0.7045	0.0090	0.6890	0.0446	0.0173	31	0.35	128.00	0
128	Isoleucine (%)	0227	0.7050	0.0100	0.6890	0.0446	0.0173	31	0.36	128.00	0
128	Isoleucine (%)	0571	0.7065	0.0030	0.6890	0.0446	0.0173	31	0.39	128.00	0
128	Isoleucine (%)	2059	0.7170	0.0080	0.6890	0.0446	0.0173	31	0.63	128.00	0
128	Isoleucine (%)	2188	0.7185	0.0130	0.6890	0.0446	0.0173	31	0.66	128.05	0
128	Isoleucine (%)	0148	0.7200	0.0000	0.6890	0.0446	0.0173	31	0.69	128.05	0
128	Isoleucine (%)	0872	0.7225	0.0010	0.6890	0.0446	0.0173	31	0.75	128.00	0
128	Isoleucine (%)	0644	0.7470	0.0080	0.6890	0.0446	0.0173	31	1.30	128.00	0
128	Isoleucine (%)	0242	0.7700	0.0200	0.6890	0.0446	0.0173	31	1.81	128.05	0
128	Isoleucine (%)	0407	0.7820	0.0260	0.6890	0.0446	0.0173	31	2.08	128.05	0
128	Isoleucine (%)	0889	0.9175	0.0250	0.6890	0.0446	0.0173	31	5.12	128.99	0
129	Leucine (%)	0208	1.495	0.0100	1.595	0.0653	0.0231	29	-1.54	129.00	0
129	Leucine (%)	0910	1.495	0.0700	1.595	0.0653	0.0231	29	-1.54	129.00	0
129	Leucine (%)	0870	1.507	0.0084	1.595	0.0653	0.0231	29	-1.36	129.00	0
129	Leucine (%)	0148	1.520	0.0000	1.595	0.0653	0.0231	29	-1.15	129.05	0
129	Leucine (%)	0675	1.535	0.0300	1.595	0.0653	0.0231	29	-0.93	129.00	0
129	Leucine (%)	0918	1.540	0.0012	1.595	0.0653	0.0231	29	-0.84	129.00	0
129	Leucine (%)	0610	1.545	0.0100	1.595	0.0653	0.0231	29	-0.77	129.05	0
129	Leucine (%)	0652	1.555	0.0100	1.595	0.0653	0.0231	29	-0.62	129.00	0
129	Leucine (%)	2196	1.557	0.0000	1.595	0.0653	0.0231	29	-0.59	129.05	0
129	Leucine (%)	0684	1.581	0.0310	1.595	0.0653	0.0231	29	-0.23	129.00	0
129	Leucine (%)	0571	1.583	0.0160	1.595	0.0653	0.0231	29	-0.19	129.00	0
129	Leucine (%)	0682	1.588	0.0000	1.595	0.0653	0.0231	29	-0.11	129.00	0
129	Leucine (%)	0626	1.589	0.0210	1.595	0.0653	0.0231	29	-0.11	129.05	0
129	Leucine (%)	2059	1.591	0.0100	1.595	0.0653	0.0231	29	-0.07	129.00	0
129	Leucine (%)	0676	1.595	0.0700	1.595	0.0653	0.0231	29	-0.01	129.05	0
129	Leucine (%)	0353	1.595	0.0300	1.595	0.0653	0.0231	29	-0.01	129.00	0
129	Leucine (%)	0098	1.609	0.0280	1.595	0.0653	0.0231	29	0.21	129.02	0
129	Leucine (%)	0968	1.610	0.0200	1.595	0.0653	0.0231	29	0.22	129.00	0
129	Leucine (%)	0859	1.613	0.0124	1.595	0.0653	0.0231	29	0.26	129.00	0
129	Leucine (%)	0878	1.613	0.0320	1.595	0.0653	0.0231	29	0.27	129.00	0
129	Leucine (%)	0644	1.623	0.0070	1.595	0.0653	0.0231	29	0.41	129.00	0
129	Leucine (%)	0227	1.630	0.0200	1.595	0.0653	0.0231	29	0.53	129.00	0
129	Leucine (%)	0872	1.631	0.0010	1.595	0.0653	0.0231	29	0.54	129.00	0
129	Leucine (%)	0171	1.647	0.0440	1.595	0.0653	0.0231	29	0.79	129.00	0
129	Leucine (%)	0354	1.648	0.0160	1.595	0.0653	0.0231	29	0.80	129.00	0
129	Leucine (%)	0242	1.690	0.0200	1.595	0.0653	0.0231	29	1.45	129.05	0
129	Leucine (%)	0619	1.695	0.0500	1.595	0.0653	0.0231	29	1.52	129.00	0
129	Leucine (%)	2188	1.714	0.0420	1.595	0.0653	0.0231	29	1.81	129.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
129	Leucine (%)	0407	1.744	0.0600	1.595	0.0653	0.0231	29	2.27	129.05	0
129	Leucine (%)	0226	1.715	0.1251	1.595	0.0653	0.0231	29	1.83	129.00	1
129	Leucine (%)	0889	0.4350	0.0000	1.595	0.0653	0.0231	29	-17.77	129.99	2
130	L-Lysine (%)	0889	0.6950	0.0000	0.8932	0.0513	0.0140	36	-3.86	130.99	0
130	L-Lysine (%)	0148	0.7950	0.0100	0.8932	0.0513	0.0140	36	-1.91	130.05	0
130	L-Lysine (%)	0675	0.8250	0.0100	0.8932	0.0513	0.0140	36	-1.33	130.00	0
130	L-Lysine (%)	0939	0.8300	0.0200	0.8932	0.0513	0.0140	36	-1.23	130.00	0
130	L-Lysine (%)	0208	0.8400	0.0000	0.8932	0.0513	0.0140	36	-1.04	130.00	0
130	L-Lysine (%)	2053	0.8425	0.0190	0.8932	0.0513	0.0140	36	-0.99	130.99	0
130	L-Lysine (%)	0910	0.8450	0.0100	0.8932	0.0513	0.0140	36	-0.94	130.00	0
130	L-Lysine (%)	0676	0.8500	0.0200	0.8932	0.0513	0.0140	36	-0.84	130.05	0
130	L-Lysine (%)	0918	0.8519	0.0056	0.8932	0.0513	0.0140	36	-0.80	130.00	0
130	L-Lysine (%)	0626	0.8550	0.0080	0.8932	0.0513	0.0140	36	-0.74	130.05	0
130	L-Lysine (%)	0512	0.8572	0.0202	0.8932	0.0513	0.0140	36	-0.70	130.00	0
130	L-Lysine (%)	0610	0.8600	0.0400	0.8932	0.0513	0.0140	36	-0.65	130.05	0
130	L-Lysine (%)	0652	0.8800	0.0200	0.8932	0.0513	0.0140	36	-0.26	130.00	0
130	L-Lysine (%)	0354	0.8860	0.0040	0.8932	0.0513	0.0140	36	-0.14	130.00	0
130	L-Lysine (%)	2196	0.8860	0.0000	0.8932	0.0513	0.0140	36	-0.14	130.05	0
130	L-Lysine (%)	0870	0.8895	0.0056	0.8932	0.0513	0.0140	36	-0.07	130.00	0
130	L-Lysine (%)	0353	0.8900	0.0200	0.8932	0.0513	0.0140	36	-0.06	130.00	0
130	L-Lysine (%)	0571	0.8930	0.0060	0.8932	0.0513	0.0140	36	0.00	130.00	0
130	L-Lysine (%)	0682	0.9020	0.0000	0.8932	0.0513	0.0140	36	0.17	130.00	0
130	L-Lysine (%)	0684	0.9030	0.0300	0.8932	0.0513	0.0140	36	0.19	130.00	0
130	L-Lysine (%)	2059	0.9060	0.0080	0.8932	0.0513	0.0140	36	0.25	130.00	0
130	L-Lysine (%)	0171	0.9075	0.0270	0.8932	0.0513	0.0140	36	0.28	130.00	0
130	L-Lysine (%)	0859	0.9082	0.0101	0.8932	0.0513	0.0140	36	0.29	130.00	0
130	L-Lysine (%)	0644	0.9100	0.0040	0.8932	0.0513	0.0140	36	0.33	130.00	0
130	L-Lysine (%)	0848	0.9100	0.0000	0.8932	0.0513	0.0140	36	0.33	130.00	0
130	L-Lysine (%)	0878	0.9140	0.0080	0.8932	0.0513	0.0140	36	0.41	130.00	0
130	L-Lysine (%)	0872	0.9185	0.0050	0.8932	0.0513	0.0140	36	0.49	130.00	0
130	L-Lysine (%)	0016	0.9240	0.0140	0.8932	0.0513	0.0140	36	0.60	130.99	0
130	L-Lysine (%)	0619	0.9265	0.0390	0.8932	0.0513	0.0140	36	0.65	130.00	0
130	L-Lysine (%)	0968	0.9350	0.0100	0.8932	0.0513	0.0140	36	0.82	130.00	0
130	L-Lysine (%)	0098	0.9440	0.0360	0.8932	0.0513	0.0140	36	0.99	130.02	0
130	L-Lysine (%)	2188	0.9565	0.0110	0.8932	0.0513	0.0140	36	1.23	130.05	0
130	L-Lysine (%)	0242	0.9650	0.0100	0.8932	0.0513	0.0140	36	1.40	130.05	0
130	L-Lysine (%)	0014	0.9760	0.0300	0.8932	0.0513	0.0140	36	1.61	130.05	0
130	L-Lysine (%)	0407	1.019	0.0340	0.8932	0.0513	0.0140	36	2.45	130.05	0
130	L-Lysine (%)	0227	1.045	0.0100	0.8932	0.0513	0.0140	36	2.96	130.00	0
130	L-Lysine (%)	0226	0.9557	0.0925	0.8932	0.0513	0.0140	36	1.22	130.00	1
131	Methionine (%)	2196	0.3020	0.0000	0.3783	0.0309	0.0102	36	-2.47	131.05	0
131	Methionine (%)	0626	0.3260	0.0200	0.3783	0.0309	0.0102	36	-1.69	131.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
131	Methionine (%)	2188	0.3290	0.0060	0.3783	0.0309	0.0102	36	-1.60	131.05	0
131	Methionine (%)	0675	0.3300	0.0200	0.3783	0.0309	0.0102	36	-1.56	131.00	0
131	Methionine (%)	0148	0.3300	0.0000	0.3783	0.0309	0.0102	36	-1.56	131.05	0
131	Methionine (%)	0016	0.3505	0.0030	0.3783	0.0309	0.0102	36	-0.90	131.05	0
131	Methionine (%)	0910	0.3550	0.0100	0.3783	0.0309	0.0102	36	-0.75	131.00	0
131	Methionine (%)	0939	0.3550	0.0100	0.3783	0.0309	0.0102	36	-0.75	131.00	0
131	Methionine (%)	0610	0.3550	0.0100	0.3783	0.0309	0.0102	36	-0.75	131.05	0
131	Methionine (%)	0619	0.3685	0.0110	0.3783	0.0309	0.0102	36	-0.32	131.00	0
131	Methionine (%)	0098	0.3685	0.0030	0.3783	0.0309	0.0102	36	-0.32	131.02	0
131	Methionine (%)	0918	0.3700	0.0000	0.3783	0.0309	0.0102	36	-0.27	131.00	0
131	Methionine (%)	0870	0.3706	0.0037	0.3783	0.0309	0.0102	36	-0.25	131.00	0
131	Methionine (%)	0354	0.3720	0.0060	0.3783	0.0309	0.0102	36	-0.20	131.00	0
131	Methionine (%)	0652	0.3750	0.0100	0.3783	0.0309	0.0102	36	-0.11	131.00	0
131	Methionine (%)	0512	0.3753	0.0255	0.3783	0.0309	0.0102	36	-0.10	131.00	0
131	Methionine (%)	0208	0.3760	0.0100	0.3783	0.0309	0.0102	36	-0.07	131.00	0
131	Methionine (%)	2059	0.3775	0.0050	0.3783	0.0309	0.0102	36	-0.03	131.00	0
131	Methionine (%)	0684	0.3790	0.0420	0.3783	0.0309	0.0102	36	0.02	131.00	0
131	Methionine (%)	0859	0.3834	0.0004	0.3783	0.0309	0.0102	36	0.17	131.00	0
131	Methionine (%)	0226	0.3843	0.0114	0.3783	0.0309	0.0102	36	0.19	131.00	0
131	Methionine (%)	0571	0.3845	0.0010	0.3783	0.0309	0.0102	36	0.20	131.00	0
131	Methionine (%)	0227	0.3850	0.0100	0.3783	0.0309	0.0102	36	0.22	131.00	0
131	Methionine (%)	0968	0.3865	0.0050	0.3783	0.0309	0.0102	36	0.27	131.00	0
131	Methionine (%)	0014	0.3885	0.0290	0.3783	0.0309	0.0102	36	0.33	131.05	0
131	Methionine (%)	0644	0.3930	0.0000	0.3783	0.0309	0.0102	36	0.48	131.00	0
131	Methionine (%)	0872	0.3955	0.0070	0.3783	0.0309	0.0102	36	0.56	131.00	0
131	Methionine (%)	0878	0.3955	0.0150	0.3783	0.0309	0.0102	36	0.56	131.00	0
131	Methionine (%)	2053	0.3990	0.0060	0.3783	0.0309	0.0102	36	0.67	131.99	0
131	Methionine (%)	0242	0.4000	0.0200	0.3783	0.0309	0.0102	36	0.70	131.05	0
131	Methionine (%)	0682	0.4070	0.0000	0.3783	0.0309	0.0102	36	0.93	131.00	0
131	Methionine (%)	0848	0.4100	0.0000	0.3783	0.0309	0.0102	36	1.03	131.00	0
131	Methionine (%)	0171	0.4390	0.0000	0.3783	0.0309	0.0102	36	1.97	131.00	0
131	Methionine (%)	0407	0.4458	0.0272	0.3783	0.0309	0.0102	36	2.19	131.05	0
131	Methionine (%)	0353	0.4500	0.0000	0.3783	0.0309	0.0102	36	2.32	131.00	0
131	Methionine (%)	0676	0.4900	0.0400	0.3783	0.0309	0.0102	36	3.62	131.05	0
131	Methionine (%)	0889	0.8425	0.0050	0.3783	0.0309	0.0102	36	15.04	131.99	2
132	Phenylalanine (%)	0889	0.6600	0.0000	0.8540	0.0301	0.0128	30	-6.45	132.99	0
132	Phenylalanine (%)	0682	0.7860	0.0000	0.8540	0.0301	0.0128	30	-2.26	132.00	0
132	Phenylalanine (%)	0910	0.8100	0.0200	0.8540	0.0301	0.0128	30	-1.46	132.00	0
132	Phenylalanine (%)	0675	0.8150	0.0100	0.8540	0.0301	0.0128	30	-1.30	132.00	0
132	Phenylalanine (%)	0652	0.8250	0.0100	0.8540	0.0301	0.0128	30	-0.96	132.00	0
132	Phenylalanine (%)	0870	0.8269	0.0121	0.8540	0.0301	0.0128	30	-0.90	132.00	0
132	Phenylalanine (%)	0610	0.8300	0.0000	0.8540	0.0301	0.0128	30	-0.80	132.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
132	Phenylalanine (%)	0626	0.8315	0.0010	0.8540	0.0301	0.0128	30	-0.75	132.05	0
132	Phenylalanine (%)	0171	0.8400	0.0040	0.8540	0.0301	0.0128	30	-0.47	132.00	0
132	Phenylalanine (%)	0968	0.8400	0.0000	0.8540	0.0301	0.0128	30	-0.47	132.00	0
132	Phenylalanine (%)	0208	0.8450	0.0100	0.8540	0.0301	0.0128	30	-0.30	132.00	0
132	Phenylalanine (%)	0684	0.8465	0.0370	0.8540	0.0301	0.0128	30	-0.25	132.00	0
132	Phenylalanine (%)	0571	0.8500	0.0080	0.8540	0.0301	0.0128	30	-0.13	132.00	0
132	Phenylalanine (%)	0878	0.8560	0.0340	0.8540	0.0301	0.0128	30	0.07	132.00	0
132	Phenylalanine (%)	0619	0.8570	0.0300	0.8540	0.0301	0.0128	30	0.10	132.00	0
132	Phenylalanine (%)	0227	0.8600	0.0200	0.8540	0.0301	0.0128	30	0.20	132.00	0
132	Phenylalanine (%)	0676	0.8600	0.0200	0.8540	0.0301	0.0128	30	0.20	132.05	0
132	Phenylalanine (%)	0407	0.8630	0.0300	0.8540	0.0301	0.0128	30	0.30	132.05	0
132	Phenylalanine (%)	0098	0.8635	0.0270	0.8540	0.0301	0.0128	30	0.32	132.02	0
132	Phenylalanine (%)	0354	0.8655	0.0090	0.8540	0.0301	0.0128	30	0.38	132.00	0
132	Phenylalanine (%)	2059	0.8670	0.0040	0.8540	0.0301	0.0128	30	0.43	132.00	0
132	Phenylalanine (%)	0353	0.8700	0.0000	0.8540	0.0301	0.0128	30	0.53	132.00	0
132	Phenylalanine (%)	0148	0.8700	0.0200	0.8540	0.0301	0.0128	30	0.53	132.05	0
132	Phenylalanine (%)	0872	0.8705	0.0090	0.8540	0.0301	0.0128	30	0.55	132.00	0
132	Phenylalanine (%)	2196	0.8760	0.0000	0.8540	0.0301	0.0128	30	0.73	132.05	0
132	Phenylalanine (%)	0644	0.8815	0.0130	0.8540	0.0301	0.0128	30	0.91	132.00	0
132	Phenylalanine (%)	0918	0.8854	0.0108	0.8540	0.0301	0.0128	30	1.04	132.00	0
132	Phenylalanine (%)	2188	0.9120	0.0160	0.8540	0.0301	0.0128	30	1.93	132.05	0
132	Phenylalanine (%)	0859	0.9195	0.0086	0.8540	0.0301	0.0128	30	2.18	132.00	0
132	Phenylalanine (%)	0242	0.9200	0.0200	0.8540	0.0301	0.0128	30	2.19	132.05	0
132	Phenylalanine (%)	0226	0.9128	0.0785	0.8540	0.0301	0.0128	30	1.95	132.00	1
133	Proline (%)	0859	0.7360	0.0175	1.186	0.0653	0.0258	29	-6.89	133.00	0
133	Proline (%)	2059	1.101	0.0120	1.186	0.0653	0.0258	29	-1.30	133.00	0
133	Proline (%)	0910	1.115	0.0100	1.186	0.0653	0.0258	29	-1.09	133.00	0
133	Proline (%)	0870	1.125	0.0138	1.186	0.0653	0.0258	29	-0.93	133.00	0
133	Proline (%)	2196	1.128	0.0000	1.186	0.0653	0.0258	29	-0.89	133.05	0
133	Proline (%)	0610	1.130	0.0200	1.186	0.0653	0.0258	29	-0.86	133.05	0
133	Proline (%)	0682	1.136	0.0000	1.186	0.0653	0.0258	29	-0.76	133.00	0
133	Proline (%)	0208	1.140	0.0800	1.186	0.0653	0.0258	29	-0.70	133.00	0
133	Proline (%)	0354	1.153	0.0070	1.186	0.0653	0.0258	29	-0.51	133.00	0
133	Proline (%)	0148	1.155	0.0500	1.186	0.0653	0.0258	29	-0.47	133.05	0
133	Proline (%)	0171	1.162	0.0210	1.186	0.0653	0.0258	29	-0.37	133.00	0
133	Proline (%)	0878	1.174	0.0010	1.186	0.0653	0.0258	29	-0.19	133.00	0
133	Proline (%)	0227	1.175	0.0100	1.186	0.0653	0.0258	29	-0.17	133.00	0
133	Proline (%)	0918	1.177	0.0188	1.186	0.0653	0.0258	29	-0.14	133.00	0
133	Proline (%)	0571	1.183	0.0100	1.186	0.0653	0.0258	29	-0.04	133.00	0
133	Proline (%)	0652	1.185	0.0300	1.186	0.0653	0.0258	29	-0.01	133.00	0
133	Proline (%)	0626	1.185	0.0220	1.186	0.0653	0.0258	29	-0.01	133.05	0
133	Proline (%)	0872	1.187	0.0300	1.186	0.0653	0.0258	29	0.02	133.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
133	Proline (%)	0684	1.189	0.0160	1.186	0.0653	0.0258	29	0.05	133.00	0
133	Proline (%)	0676	1.200	0.0600	1.186	0.0653	0.0258	29	0.22	133.05	0
133	Proline (%)	0644	1.204	0.0270	1.186	0.0653	0.0258	29	0.27	133.00	0
133	Proline (%)	0619	1.220	0.0600	1.186	0.0653	0.0258	29	0.52	133.00	0
133	Proline (%)	0242	1.225	0.0100	1.186	0.0653	0.0258	29	0.60	133.05	0
133	Proline (%)	2188	1.247	0.0190	1.186	0.0653	0.0258	29	0.93	133.05	0
133	Proline (%)	0968	1.275	0.0300	1.186	0.0653	0.0258	29	1.36	133.00	0
133	Proline (%)	0353	1.305	0.0300	1.186	0.0653	0.0258	29	1.82	133.00	0
133	Proline (%)	0407	1.341	0.0730	1.186	0.0653	0.0258	29	2.37	133.05	0
133	Proline (%)	0675	1.355	0.0700	1.186	0.0653	0.0258	29	2.59	133.00	0
133	Proline (%)	0889	1.740	0.0000	1.186	0.0653	0.0258	29	8.48	133.99	0
133	Proline (%)	0226	1.344	0.2171	1.186	0.0653	0.0258	29	2.42	133.00	1
134	Serine (%)	0407	0.7950	0.0180	0.8696	0.0559	0.0166	31	-1.34	134.05	0
134	Serine (%)	0610	0.7950	0.0300	0.8696	0.0559	0.0166	31	-1.34	134.05	0
134	Serine (%)	0148	0.8000	0.0400	0.8696	0.0559	0.0166	31	-1.25	134.05	0
134	Serine (%)	0171	0.8065	0.0190	0.8696	0.0559	0.0166	31	-1.13	134.00	0
134	Serine (%)	0675	0.8150	0.0100	0.8696	0.0559	0.0166	31	-0.98	134.00	0
134	Serine (%)	0644	0.8225	0.0110	0.8696	0.0559	0.0166	31	-0.84	134.00	0
134	Serine (%)	0208	0.8300	0.0200	0.8696	0.0559	0.0166	31	-0.71	134.00	0
134	Serine (%)	0910	0.8300	0.0200	0.8696	0.0559	0.0166	31	-0.71	134.00	0
134	Serine (%)	0098	0.8370	0.0100	0.8696	0.0559	0.0166	31	-0.58	134.02	0
134	Serine (%)	0227	0.8450	0.0100	0.8696	0.0559	0.0166	31	-0.44	134.00	0
134	Serine (%)	0571	0.8460	0.0100	0.8696	0.0559	0.0166	31	-0.42	134.00	0
134	Serine (%)	2059	0.8460	0.0000	0.8696	0.0559	0.0166	31	-0.42	134.00	0
134	Serine (%)	0872	0.8510	0.0060	0.8696	0.0559	0.0166	31	-0.33	134.00	0
134	Serine (%)	0626	0.8545	0.0110	0.8696	0.0559	0.0166	31	-0.27	134.05	0
134	Serine (%)	0918	0.8555	0.0016	0.8696	0.0559	0.0166	31	-0.25	134.00	0
134	Serine (%)	0652	0.8600	0.0200	0.8696	0.0559	0.0166	31	-0.17	134.00	0
134	Serine (%)	0968	0.8600	0.0000	0.8696	0.0559	0.0166	31	-0.17	134.00	0
134	Serine (%)	0878	0.8645	0.0310	0.8696	0.0559	0.0166	31	-0.09	134.00	0
134	Serine (%)	0870	0.8768	0.0618	0.8696	0.0559	0.0166	31	0.13	134.00	0
134	Serine (%)	2196	0.8830	0.0000	0.8696	0.0559	0.0166	31	0.24	134.05	0
134	Serine (%)	0242	0.9000	0.0000	0.8696	0.0559	0.0166	31	0.54	134.05	0
134	Serine (%)	0354	0.9010	0.0060	0.8696	0.0559	0.0166	31	0.56	134.00	0
134	Serine (%)	0682	0.9010	0.0000	0.8696	0.0559	0.0166	31	0.56	134.00	0
134	Serine (%)	2188	0.9060	0.0060	0.8696	0.0559	0.0166	31	0.65	134.05	0
134	Serine (%)	0684	0.9085	0.0610	0.8696	0.0559	0.0166	31	0.70	134.00	0
134	Serine (%)	0859	0.9251	0.0069	0.8696	0.0559	0.0166	31	0.99	134.00	0
134	Serine (%)	0353	0.9300	0.0200	0.8696	0.0559	0.0166	31	1.08	134.00	0
134	Serine (%)	0619	0.9870	0.0060	0.8696	0.0559	0.0166	31	2.10	134.00	0
134	Serine (%)	0226	0.9896	0.0506	0.8696	0.0559	0.0166	31	2.15	134.00	0
134	Serine (%)	0676	0.9900	0.0200	0.8696	0.0559	0.0166	31	2.15	134.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
134	Serine (%)	0889	1.145	0.0100	0.8696	0.0559	0.0166	31	4.93	134.99	0
135	Threonine (%)	0407	0.5220	0.0060	0.6820	0.0358	0.0104	30	-4.47	135.05	0
135	Threonine (%)	0626	0.6225	0.0050	0.6820	0.0358	0.0104	30	-1.66	135.05	0
135	Threonine (%)	0652	0.6400	0.0000	0.6820	0.0358	0.0104	30	-1.17	135.00	0
135	Threonine (%)	0610	0.6400	0.0200	0.6820	0.0358	0.0104	30	-1.17	135.05	0
135	Threonine (%)	0208	0.6450	0.0100	0.6820	0.0358	0.0104	30	-1.04	135.00	0
135	Threonine (%)	0675	0.6500	0.0000	0.6820	0.0358	0.0104	30	-0.90	135.00	0
135	Threonine (%)	0910	0.6500	0.0200	0.6820	0.0358	0.0104	30	-0.90	135.00	0
135	Threonine (%)	0148	0.6600	0.0200	0.6820	0.0358	0.0104	30	-0.62	135.05	0
135	Threonine (%)	2059	0.6630	0.0020	0.6820	0.0358	0.0104	30	-0.53	135.00	0
135	Threonine (%)	0571	0.6645	0.0090	0.6820	0.0358	0.0104	30	-0.49	135.00	0
135	Threonine (%)	0918	0.6692	0.0194	0.6820	0.0358	0.0104	30	-0.36	135.00	0
135	Threonine (%)	0354	0.6710	0.0020	0.6820	0.0358	0.0104	30	-0.31	135.00	0
135	Threonine (%)	0872	0.6750	0.0080	0.6820	0.0358	0.0104	30	-0.20	135.00	0
135	Threonine (%)	0644	0.6785	0.0090	0.6820	0.0358	0.0104	30	-0.10	135.00	0
135	Threonine (%)	0682	0.6790	0.0000	0.6820	0.0358	0.0104	30	-0.09	135.00	0
135	Threonine (%)	0870	0.6830	0.0003	0.6820	0.0358	0.0104	30	0.03	135.00	0
135	Threonine (%)	0353	0.6850	0.0100	0.6820	0.0358	0.0104	30	0.08	135.00	0
135	Threonine (%)	2196	0.6850	0.0000	0.6820	0.0358	0.0104	30	0.08	135.05	0
135	Threonine (%)	0098	0.6915	0.0070	0.6820	0.0358	0.0104	30	0.26	135.02	0
135	Threonine (%)	0878	0.6965	0.0010	0.6820	0.0358	0.0104	30	0.40	135.00	0
135	Threonine (%)	0859	0.6983	0.0062	0.6820	0.0358	0.0104	30	0.45	135.00	0
135	Threonine (%)	0171	0.7010	0.0400	0.6820	0.0358	0.0104	30	0.53	135.00	0
135	Threonine (%)	0676	0.7050	0.0100	0.6820	0.0358	0.0104	30	0.64	135.05	0
135	Threonine (%)	0227	0.7150	0.0100	0.6820	0.0358	0.0104	30	0.92	135.00	0
135	Threonine (%)	0684	0.7200	0.0340	0.6820	0.0358	0.0104	30	1.06	135.00	0
135	Threonine (%)	0968	0.7200	0.0200	0.6820	0.0358	0.0104	30	1.06	135.00	0
135	Threonine (%)	2188	0.7225	0.0130	0.6820	0.0358	0.0104	30	1.13	135.05	0
135	Threonine (%)	0242	0.7300	0.0000	0.6820	0.0358	0.0104	30	1.34	135.05	0
135	Threonine (%)	0619	0.7310	0.0100	0.6820	0.0358	0.0104	30	1.37	135.00	0
135	Threonine (%)	0889	0.8800	0.0200	0.6820	0.0358	0.0104	30	5.53	135.99	0
135	Threonine (%)	0226	0.7324	0.0567	0.6820	0.0358	0.0104	30	1.41	135.00	1
136	Tryptophan (%)	0889	0.0700	0.0000	0.1961	0.0258	0.0064	18	-4.89	136.99	0
136	Tryptophan (%)	0407	0.1140	0.0220	0.1961	0.0258	0.0064	18	-3.18	136.05	0
136	Tryptophan (%)	0918	0.1400	0.0200	0.1961	0.0258	0.0064	18	-2.17	136.00	0
136	Tryptophan (%)	0208	0.1620	0.0020	0.1961	0.0258	0.0064	18	-1.32	136.01	0
136	Tryptophan (%)	0859	0.1908	0.0109	0.1961	0.0258	0.0064	18	-0.21	136.03	0
136	Tryptophan (%)	0242	0.1950	0.0100	0.1961	0.0258	0.0064	18	-0.04	136.05	0
136	Tryptophan (%)	0682	0.1960	0.0000	0.1961	0.0258	0.0064	18	0.00	136.00	0
136	Tryptophan (%)	0619	0.2005	0.0090	0.1961	0.0258	0.0064	18	0.17	136.00	0
136	Tryptophan (%)	0878	0.2015	0.0030	0.1961	0.0258	0.0064	18	0.21	136.01	0
136	Tryptophan (%)	0872	0.2015	0.0030	0.1961	0.0258	0.0064	18	0.21	136.03	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
136	Tryptophan (%)	0098	0.2030	0.0100	0.1961	0.0258	0.0064	18	0.27	136.02	0
136	Tryptophan (%)	2059	0.2065	0.0030	0.1961	0.0258	0.0064	18	0.40	136.03	0
136	Tryptophan (%)	0644	0.2080	0.0040	0.1961	0.0258	0.0064	18	0.46	136.01	0
136	Tryptophan (%)	0571	0.2110	0.0020	0.1961	0.0258	0.0064	18	0.58	136.03	0
136	Tryptophan (%)	0870	0.2132	0.0047	0.1961	0.0258	0.0064	18	0.66	136.00	0
136	Tryptophan (%)	0684	0.2185	0.0110	0.1961	0.0258	0.0064	18	0.87	136.00	0
136	Tryptophan (%)	0353	0.2200	0.0000	0.1961	0.0258	0.0064	18	0.93	136.03	0
136	Tryptophan (%)	0227	0.2300	0.0000	0.1961	0.0258	0.0064	18	1.31	136.00	0
136	Tryptophan (%)	0910	0.1800	0.0400	0.1961	0.0258	0.0064	18	-0.62	136.00	1
137	Tyrosine (%)	0889	0.0700	0.0000	0.6233	0.0657	0.0167	26	-8.42	137.99	0
137	Tyrosine (%)	0910	0.5100	0.0400	0.6233	0.0657	0.0167	26	-1.73	137.00	0
137	Tyrosine (%)	0208	0.5250	0.0300	0.6233	0.0657	0.0167	26	-1.50	137.00	0
137	Tyrosine (%)	0227	0.5350	0.0100	0.6233	0.0657	0.0167	26	-1.34	137.01	0
137	Tyrosine (%)	2196	0.5730	0.0000	0.6233	0.0657	0.0167	26	-0.77	137.05	0
137	Tyrosine (%)	0626	0.5800	0.0000	0.6233	0.0657	0.0167	26	-0.66	137.05	0
137	Tyrosine (%)	0968	0.5850	0.0100	0.6233	0.0657	0.0167	26	-0.58	137.00	0
137	Tyrosine (%)	0098	0.6060	0.0300	0.6233	0.0657	0.0167	26	-0.26	137.02	0
137	Tyrosine (%)	0354	0.6085	0.0050	0.6233	0.0657	0.0167	26	-0.23	137.00	0
137	Tyrosine (%)	0872	0.6095	0.0050	0.6233	0.0657	0.0167	26	-0.21	137.00	0
137	Tyrosine (%)	0918	0.6183	0.0280	0.6233	0.0657	0.0167	26	-0.08	137.00	0
137	Tyrosine (%)	0226	0.6205	0.0573	0.6233	0.0657	0.0167	26	-0.04	137.00	0
137	Tyrosine (%)	0682	0.6210	0.0000	0.6233	0.0657	0.0167	26	-0.04	137.00	0
137	Tyrosine (%)	0171	0.6240	0.0300	0.6233	0.0657	0.0167	26	0.01	137.00	0
137	Tyrosine (%)	0684	0.6260	0.0000	0.6233	0.0657	0.0167	26	0.04	137.00	0
137	Tyrosine (%)	0870	0.6316	0.0158	0.6233	0.0657	0.0167	26	0.13	137.00	0
137	Tyrosine (%)	0644	0.6335	0.0110	0.6233	0.0657	0.0167	26	0.15	137.00	0
137	Tyrosine (%)	0242	0.6450	0.0100	0.6233	0.0657	0.0167	26	0.33	137.05	0
137	Tyrosine (%)	0675	0.6500	0.0000	0.6233	0.0657	0.0167	26	0.41	137.00	0
137	Tyrosine (%)	0610	0.6650	0.0300	0.6233	0.0657	0.0167	26	0.63	137.05	0
137	Tyrosine (%)	2188	0.6720	0.0180	0.6233	0.0657	0.0167	26	0.74	137.05	0
137	Tyrosine (%)	2059	0.6895	0.0010	0.6233	0.0657	0.0167	26	1.01	137.00	0
137	Tyrosine (%)	0353	0.6900	0.0000	0.6233	0.0657	0.0167	26	1.01	137.00	0
137	Tyrosine (%)	0676	0.7050	0.0300	0.6233	0.0657	0.0167	26	1.24	137.05	0
137	Tyrosine (%)	0148	0.7250	0.0300	0.6233	0.0657	0.0167	26	1.55	137.05	0
137	Tyrosine (%)	0407	1.273	0.0420	0.6233	0.0657	0.0167	26	9.89	137.05	0
138	Valine (%)	0889	0.3400	0.0000	0.8110	0.0595	0.0184	31	-7.92	138.99	0
138	Valine (%)	0226	0.6227	0.0622	0.8110	0.0595	0.0184	31	-3.16	138.00	0
138	Valine (%)	0675	0.6450	0.0300	0.8110	0.0595	0.0184	31	-2.79	138.00	0
138	Valine (%)	0171	0.7225	0.0470	0.8110	0.0595	0.0184	31	-1.49	138.00	0
138	Valine (%)	0353	0.7350	0.0100	0.8110	0.0595	0.0184	31	-1.28	138.00	0
138	Valine (%)	0619	0.7415	0.0570	0.8110	0.0595	0.0184	31	-1.17	138.00	0
138	Valine (%)	0910	0.7750	0.0500	0.8110	0.0595	0.0184	31	-0.60	138.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
138	Valine (%)	0968	0.7800	0.0000	0.8110	0.0595	0.0184	31	-0.52	138.00	0
138	Valine (%)	0870	0.7810	0.0414	0.8110	0.0595	0.0184	31	-0.50	138.00	0
138	Valine (%)	0918	0.7851	0.0254	0.8110	0.0595	0.0184	31	-0.43	138.00	0
138	Valine (%)	0652	0.7950	0.0500	0.8110	0.0595	0.0184	31	-0.27	138.00	0
138	Valine (%)	2196	0.8010	0.0000	0.8110	0.0595	0.0184	31	-0.17	138.05	0
138	Valine (%)	0859	0.8084	0.0071	0.8110	0.0595	0.0184	31	-0.04	138.00	0
138	Valine (%)	0626	0.8110	0.0000	0.8110	0.0595	0.0184	31	0.00	138.05	0
138	Valine (%)	0208	0.8250	0.0300	0.8110	0.0595	0.0184	31	0.24	138.00	0
138	Valine (%)	0684	0.8250	0.0060	0.8110	0.0595	0.0184	31	0.24	138.00	0
138	Valine (%)	0878	0.8250	0.0140	0.8110	0.0595	0.0184	31	0.24	138.00	0
138	Valine (%)	0682	0.8270	0.0000	0.8110	0.0595	0.0184	31	0.27	138.00	0
138	Valine (%)	0571	0.8280	0.0080	0.8110	0.0595	0.0184	31	0.29	138.00	0
138	Valine (%)	2059	0.8300	0.0040	0.8110	0.0595	0.0184	31	0.32	138.00	0
138	Valine (%)	0148	0.8350	0.0100	0.8110	0.0595	0.0184	31	0.40	138.05	0
138	Valine (%)	0354	0.8370	0.0040	0.8110	0.0595	0.0184	31	0.44	138.00	0
138	Valine (%)	0872	0.8385	0.0050	0.8110	0.0595	0.0184	31	0.46	138.00	0
138	Valine (%)	0676	0.8400	0.0200	0.8110	0.0595	0.0184	31	0.49	138.05	0
138	Valine (%)	0610	0.8500	0.0200	0.8110	0.0595	0.0184	31	0.66	138.05	0
138	Valine (%)	0098	0.8505	0.0070	0.8110	0.0595	0.0184	31	0.66	138.02	0
138	Valine (%)	0227	0.8650	0.0100	0.8110	0.0595	0.0184	31	0.91	138.00	0
138	Valine (%)	0644	0.8830	0.0060	0.8110	0.0595	0.0184	31	1.21	138.00	0
138	Valine (%)	2188	0.8850	0.0140	0.8110	0.0595	0.0184	31	1.24	138.05	0
138	Valine (%)	0242	0.8950	0.0100	0.8110	0.0595	0.0184	31	1.41	138.05	0
138	Valine (%)	0407	0.9325	0.0210	0.8110	0.0595	0.0184	31	2.04	138.05	0
139	Taurine (%)	0407	0.0001	0.0000				3		139.05	0
139	Taurine (%)	0889	0.0100	0.0000				3		139.99	0
139	Taurine (%)	0682	0.0820	0.0000				3		139.00	0
139	Taurine (%)	0242	0.0000	0.0000				3		139.05	4
139	Taurine (%)	0227	< 0.01					3		139.00	5
139	Taurine (%)	0098	< 0.01					3		139.02	5
160	Fructose (%)	0297	0.0905	0.0190				2		160.10	0
160	Fructose (%)	0619	0.0905	0.0010				2		160.99	0
160	Fructose (%)	0227	< 0.15					2		160.99	5
161	Galactose (%)	0297	0.0000	0.0000				0		161.10	4
162	Glucose (%)	0227	0.1850	0.0300				3		162.99	0
162	Glucose (%)	0297	0.2565	0.0150				3		162.10	0
162	Glucose (%)	0619	0.2585	0.0010				3		162.99	0
163	Lactose (%)	0297	0.0000	0.0000				0		163.10	4
163	Lactose (%)	0619	0.0000	0.0000				0		163.99	4
163	Lactose (%)	0227	< 0.15					0		163.99	5
164	Maltose (%)	0297	0.2040	0.0020				1		164.10	0
164	Maltose (%)	0227	< 0.15					1		164.99	5

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
165	Sucrose (%)	0619	1.560	0.0000				3		165.99	0
165	Sucrose (%)	0227	1.750	0.0200				3		165.99	0
165	Sucrose (%)	0297	1.937	0.1890				3		165.10	0
166	Raffinose (%)	0619	0.2805	0.0030				3		166.99	0
166	Raffinose (%)	0227	0.2850	0.0100				3		166.99	0
166	Raffinose (%)	0297	0.3395	0.0110				3		166.10	0
167	Stachyose (%)	0227	0.9100	0.0200				3		167.99	0
167	Stachyose (%)	0619	0.9350	0.0040				3		167.99	0
167	Stachyose (%)	0297	0.9565	0.0510				3		167.10	0
345	Amprolium (ppm)	0875	41.00	4.000	63.35	14.13	3.472	15	-1.58	345.04	0
345	Amprolium (ppm)	0964	53.72	1.807	63.35	14.13	3.472	15	-0.68	345.02	0
345	Amprolium (ppm)	0014	55.45	3.500	63.35	14.13	3.472	15	-0.56	345.02	0
345	Amprolium (ppm)	0032	55.50	3.000	63.35	14.13	3.472	15	-0.56	345.02	0
345	Amprolium (ppm)	0036	55.80	2.400	63.35	14.13	3.472	15	-0.53	345.00	0
345	Amprolium (ppm)	0512	57.49	0.1000	63.35	14.13	3.472	15	-0.41	345.02	0
345	Amprolium (ppm)	0218	58.01	1.010	63.35	14.13	3.472	15	-0.38	345.02	0
345	Amprolium (ppm)	0171	60.40	0.4000	63.35	14.13	3.472	15	-0.21	345.00	0
345	Amprolium (ppm)	2192	61.68	2.010	63.35	14.13	3.472	15	-0.12	345.02	0
345	Amprolium (ppm)	0511	61.91	1.660	63.35	14.13	3.472	15	-0.10	345.00	0
345	Amprolium (ppm)	0016	62.65	12.10	63.35	14.13	3.472	15	-0.05	345.00	0
345	Amprolium (ppm)	0148	74.08	9.450	63.35	14.13	3.472	15	0.76	345.04	0
345	Amprolium (ppm)	0028	82.32	0.4700	63.35	14.13	3.472	15	1.34	345.00	0
345	Amprolium (ppm)	0017	85.50	1.000	63.35	14.13	3.472	15	1.57	345.00	0
345	Amprolium (ppm)	2234	116.0	9.180	63.35	14.13	3.472	15	3.73	345.04	0
379	Salinomycin (ppm)	2113	0.4300	0.0000				1		379.05	0
391	Narasin (ppm)	0148	99.36	1.030	126.4	14.83	4.400	15	-1.82	391.03	0
391	Narasin (ppm)	2234	105.9	4.350	126.4	14.83	4.400	15	-1.38	391.03	0
391	Narasin (ppm)	0619	107.2	8.155	126.4	14.83	4.400	15	-1.29	391.03	0
391	Narasin (ppm)	0875	119.0	6.000	126.4	14.83	4.400	15	-0.50	391.03	0
391	Narasin (ppm)	0668	121.4	4.500	126.4	14.83	4.400	15	-0.34	391.00	0
391	Narasin (ppm)	0047	123.7	0.3000	126.4	14.83	4.400	15	-0.19	391.01	0
391	Narasin (ppm)	0033	127.0	12.00	126.4	14.83	4.400	15	0.04	391.00	0
391	Narasin (ppm)	2053	127.0	0.0000	126.4	14.83	4.400	15	0.04	391.03	0
391	Narasin (ppm)	0148	129.6	3.130	126.4	14.83	4.400	15	0.21	391.00	0
391	Narasin (ppm)	0910	132.3	1.200	126.4	14.83	4.400	15	0.40	391.99	0
391	Narasin (ppm)	0218	134.3	2.835	126.4	14.83	4.400	15	0.53	391.00	0
391	Narasin (ppm)	0668	135.5	1.000	126.4	14.83	4.400	15	0.61	391.99	0
391	Narasin (ppm)	0610	142.0	6.000	126.4	14.83	4.400	15	1.05	391.01	0
391	Narasin (ppm)	0610	143.5	3.000	126.4	14.83	4.400	15	1.15	391.03	0
391	Narasin (ppm)	2113	143.8	12.50	126.4	14.83	4.400	15	1.17	391.03	0
400	Water Activity (Units)	0843	0.5580	0.0000	0.5804	0.0202	0.0040	10	-1.11	400.99	0
400	Water Activity (Units)	2303	0.5597	0.0129	0.5804	0.0202	0.0040	10	-1.03	400.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
400	Water Activity (Units)	0942	0.5680	0.0020	0.5804	0.0202	0.0040	10	-0.61	400.01	0
400	Water Activity (Units)	0843	0.5699	0.0161	0.5804	0.0202	0.0040	10	-0.52	400.01	0
400	Water Activity (Units)	0407	0.5767	0.0012	0.5804	0.0202	0.0040	10	-0.18	400.01	0
400	Water Activity (Units)	2033	0.5800	0.0000	0.5804	0.0202	0.0040	10	-0.02	400.99	0
400	Water Activity (Units)	0589	0.5825	0.0030	0.5804	0.0202	0.0040	10	0.10	400.01	0
400	Water Activity (Units)	2181	0.5965	0.0013	0.5804	0.0202	0.0040	10	0.80	400.01	0
400	Water Activity (Units)	2073	0.6020	0.0020	0.5804	0.0202	0.0040	10	1.07	400.99	0
400	Water Activity (Units)	2109	0.6315	0.0015	0.5804	0.0202	0.0040	10	2.53	400.01	0
400	Water Activity (Units)	2168	0.5735	0.0330	0.5804	0.0202	0.0040	10	-0.34	400.01	1
516	Arsenic, Total (ppm)	0045	0.0990	0.0040	0.1314	0.0312	0.0144	6	-1.04	516.00	0
516	Arsenic, Total (ppm)	0186	0.1120	0.0000	0.1314	0.0312	0.0144	6	-0.62	516.52	0
516	Arsenic, Total (ppm)	0227	0.1235	0.0150	0.1314	0.0312	0.0144	6	-0.25	516.53	0
516	Arsenic, Total (ppm)	0910	0.1250	0.0100	0.1314	0.0312	0.0144	6	-0.21	516.52	0
516	Arsenic, Total (ppm)	0918	0.1589	0.0173	0.1314	0.0312	0.0144	6	0.88	516.53	0
516	Arsenic, Total (ppm)	2053	0.1700	0.0400	0.1314	0.0312	0.0144	6	1.24	516.53	0
516	Arsenic, Total (ppm)	0407	1.772	0.1250	0.1314	0.0312	0.0144	6	52.65	516.43	1
516	Arsenic, Total (ppm)	0682	< 10		0.1314	0.0312	0.0144	6		516.43	5
518	Cadmium (ppm)	0171	0.0545	0.0010	0.0651	0.0084	0.0023	10	-1.26	518.41	0
518	Cadmium (ppm)	0939	0.0550	0.0000	0.0651	0.0084	0.0023	10	-1.20	518.53	0
518	Cadmium (ppm)	0968	0.0605	0.0030	0.0651	0.0084	0.0023	10	-0.55	518.43	0
518	Cadmium (ppm)	0407	0.0623	0.0063	0.0651	0.0084	0.0023	10	-0.34	518.41	0
518	Cadmium (ppm)	0227	0.0650	0.0000	0.0651	0.0084	0.0023	10	-0.01	518.53	0
518	Cadmium (ppm)	0186	0.0655	0.0010	0.0651	0.0084	0.0023	10	0.04	518.52	0
518	Cadmium (ppm)	0918	0.0677	0.0018	0.0651	0.0084	0.0023	10	0.31	518.53	0
518	Cadmium (ppm)	0910	0.0700	0.0000	0.0651	0.0084	0.0023	10	0.58	518.52	0
518	Cadmium (ppm)	0407	0.0730	0.0068	0.0651	0.0084	0.0023	10	0.93	518.43	0
518	Cadmium (ppm)	0045	0.2065	0.0030	0.0651	0.0084	0.0023	10	16.77	518.42	0
518	Cadmium (ppm)	0689	< 0.2		0.0651	0.0084	0.0023	10		518.31	5
518	Cadmium (ppm)	0682	< 0.2		0.0651	0.0084	0.0023	10		518.43	5
520	Chromium (ppm)	0407	1.094	0.2303	2.261	1.211	0.0749	13	-0.96	520.41	0
520	Chromium (ppm)	0186	1.124	0.0080	2.261	1.211	0.0749	13	-0.94	520.52	0
520	Chromium (ppm)	0407	1.156	0.0643	2.261	1.211	0.0749	13	-0.91	520.43	0
520	Chromium (ppm)	0968	1.253	0.0480	2.261	1.211	0.0749	13	-0.83	520.43	0
520	Chromium (ppm)	0918	1.272	0.0149	2.261	1.211	0.0749	13	-0.82	520.53	0
520	Chromium (ppm)	0171	1.354	0.0440	2.261	1.211	0.0749	13	-0.75	520.41	0
520	Chromium (ppm)	0693	2.669	0.1740	2.261	1.211	0.0749	13	0.34	520.42	0
520	Chromium (ppm)	0682	2.750	0.0000	2.261	1.211	0.0749	13	0.40	520.43	0
520	Chromium (ppm)	0164	2.955	0.1700	2.261	1.211	0.0749	13	0.57	520.53	0
520	Chromium (ppm)	0227	3.075	0.1100	2.261	1.211	0.0749	13	0.67	520.53	0
520	Chromium (ppm)	0045	3.125	0.0100	2.261	1.211	0.0749	13	0.71	520.42	0
520	Chromium (ppm)	0510	3.610	0.0600	2.261	1.211	0.0749	13	1.11	520.43	0
520	Chromium (ppm)	0939	3.960	0.0400	2.261	1.211	0.0749	13	1.40	520.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
526	Lead (ppm)	0171	0.1210	0.0020	0.2473	0.0749	0.0275	9	-1.69	526.41	0
526	Lead (ppm)	0407	0.2041	0.0510	0.2473	0.0749	0.0275	9	-0.58	526.41	0
526	Lead (ppm)	0910	0.2250	0.0100	0.2473	0.0749	0.0275	9	-0.30	526.52	0
526	Lead (ppm)	0939	0.2300	0.0800	0.2473	0.0749	0.0275	9	-0.23	526.53	0
526	Lead (ppm)	0227	0.2350	0.0000	0.2473	0.0749	0.0275	9	-0.16	526.53	0
526	Lead (ppm)	0918	0.2504	0.0067	0.2473	0.0749	0.0275	9	0.04	526.53	0
526	Lead (ppm)	2053	0.2550	0.0100	0.2473	0.0749	0.0275	9	0.10	526.53	0
526	Lead (ppm)	0407	0.3317	0.0775	0.2473	0.0749	0.0275	9	1.13	526.43	0
526	Lead (ppm)	0689	2.195	0.0100	0.2473	0.0749	0.0275	9	26.01	526.31	0
526	Lead (ppm)	0682	< 3		0.2473	0.0749	0.0275	9		526.43	5
529	Mercury (ppb)	0227	< 10					0		529.99	5
529	Mercury (ppb)	0682	< 10					0		529.99	5
539	Nickel (ppm)	0407	1.381	0.0911	2.202	0.8570	0.0795	7	-0.96	539.43	0
539	Nickel (ppm)	0171	1.674	0.0350	2.202	0.8570	0.0795	7	-0.62	539.41	0
539	Nickel (ppm)	0407	1.746	0.1806	2.202	0.8570	0.0795	7	-0.53	539.41	0
539	Nickel (ppm)	0186	1.759	0.0670	2.202	0.8570	0.0795	7	-0.52	539.52	0
539	Nickel (ppm)	0939	2.460	0.1200	2.202	0.8570	0.0795	7	0.30	539.53	0
539	Nickel (ppm)	0918	3.094	0.0625	2.202	0.8570	0.0795	7	1.04	539.53	0
539	Nickel (ppm)	0682	3.300	0.0000	2.202	0.8570	0.0795	7	1.28	539.43	0
702	Butyric Acid (4:0) (%)	0619	0.0000	0.0000				0		702.00	4
704	Caproic Acid (6:0) (%)	0619	0.0000	0.0000				0		704.00	4
706	Caprylic acid (8:0) (%) (w/w)	0619	0.0000	0.0000				0		706.01	4
708	Capric acid (10:0) (%) (w/w)	0619	0.0000	0.0000				0		708.01	4
710	Lauric Acid (12:0) (%) (w/w)	0676	0.0020	0.0000				1		710.99	0
710	Lauric Acid (12:0) (%) (w/w)	0619	0.0000	0.0000				1		710.01	4
710	Lauric Acid (12:0) (%) (w/w)	0226	< 0.005					1		710.99	5
710	Lauric Acid (12:0) (%) (w/w)	0164	< 0.02					1		710.99	5
714	Myristic Acid (14:0) (%) (w/w)	0619	0.0000	0.0000				0		714.01	4
714	Myristic Acid (14:0) (%) (w/w)	0226	< 0.005					0		714.99	5
714	Myristic Acid (14:0) (%) (w/w)	0164	< 0.02					0		714.99	5
716	Palmitic Acid (16:0) (%) (w/w)	0619	0.5455	0.0130				3		716.01	0
716	Palmitic Acid (16:0) (%) (w/w)	0226	0.5665	0.0070				3		716.99	0
716	Palmitic Acid (16:0) (%) (w/w)	0164	0.6000	0.0000				3		716.99	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0226	0.0050	0.0000				2		718.99	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0676	0.0055	0.0010				2		718.99	0
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0619	0.0000	0.0000				2		718.01	4
718	Palmitoleic Acid (9c-16:1) (%) (w/w)	0164	< 0.04					2		718.99	5
722	Stearic Acid (18:0) (%) (w/w)	0619	0.0815	0.0010				3		722.01	0
722	Stearic Acid (18:0) (%) (w/w)	0226	0.0975	0.0010				3		722.99	0
722	Stearic Acid (18:0) (%) (w/w)	0164	0.1000	0.0000				3		722.99	0
724	Oleic Acid (9c-18:1) (%) (w/w)	0164	1.055	0.0100				3		724.99	0
724	Oleic Acid (9c-18:1) (%) (w/w)	0226	1.071	0.0510				3		724.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
724	Oleic Acid (9c-18:1) (% (w/w))	0619	1.115	0.0500				3	724.01	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0164	1.870	0.0200	1.905	0.0349	0.0313	4	726.99	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0148	1.895	0.0300	1.905	0.0349	0.0313	4	726.02	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0226	1.901	0.0630	1.905	0.0349	0.0313	4	726.99	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0676	1.953	0.0120	1.905	0.0349	0.0313	4	726.99	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0619	2.430	0.0800	1.905	0.0349	0.0313	4	726.01	2	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v)	0619	0.3345	0.0090	0.3545	0.0175	0.0075	4	728.01	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v)	0226	0.3515	0.0090	0.3545	0.0175	0.0075	4	728.99	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v)	0164	0.3550	0.0100	0.3545	0.0175	0.0075	4	728.99	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v)	0676	0.3770	0.0020	0.3545	0.0175	0.0075	4	728.99	0	
730	Arachidic Acid (20:0) (% (w/w))	0226	0.0150	0.0000				2	730.99	0	
730	Arachidic Acid (20:0) (% (w/w))	0619	0.0290	0.0020				2	730.01	0	
730	Arachidic Acid (20:0) (% (w/w))	0164	< 0.02					2	730.99	5	
732	Gondoic Acid (11c-20:1) (% (w/w))	0226	0.0125	0.0010				2	732.99	0	
732	Gondoic Acid (11c-20:1) (% (w/w))	0619	0.0300	0.0000				2	732.01	0	
732	Gondoic Acid (11c-20:1) (% (w/w))	0164	< 0.02					2	732.99	5	
736	Arachidonic Acid (5c,8c,11c,14c-20:4) (% (w/v)	0619	0.0000	0.0000				0	736.01	4	
736	Arachidonic Acid (5c,8c,11c,14c-20:4) (% (w/v)	0676	0.0000	0.0000				0	736.99	4	
738	Mead Acid (11c,14c,17c-20:3) (% (w/w))	0619	0.0000	0.0000				0	738.01	4	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c)	0619	0.0000	0.0000				0	740.01	4	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c)	0226	0.0000	0.0000				0	740.99	4	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c)	0676	0.0000	0.0000				0	740.99	4	
742	Behenic Acid (22:0) (% (w/w))	0226	0.0090	0.0000				1	742.99	0	
742	Behenic Acid (22:0) (% (w/w))	0164	< 0.02					1	742.99	5	
744	Erucic Acid (13c-22:1) (% (w/w))	0619	0.0000	0.0000				0	744.01	4	
744	Erucic Acid (13c-22:1) (% (w/w))	0226	< 0.005					0	744.99	5	
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c)	0619	0.0000	0.0000				0	746.01	4	
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c)	0226	0.0000	0.0000				0	746.99	4	
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c)	0676	0.0000	0.0000				0	746.99	4	
748	Lignoceric Acid (24:0) (% (w/w))	0226	0.0105	0.0010				1	748.99	0	
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c)	0619	0.0000	0.0000				0	750.01	4	
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c)	0226	0.0000	0.0000				0	750.99	4	
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c)	0676	0.0000	0.0000				0	750.99	4	
752	Nervonic Acid (24:1) isomers (% (w/w))	0619	0.0000	0.0000				0	752.01	4	
752	Nervonic Acid (24:1) isomers (% (w/w))	0226	0.0000	0.0000				0	752.99	4	
754	Total n-3 Polyunsaturated (Omega-3) Fatty Acids	0164	0.3600	0.0000	0.3775	0.0171	0.0000	4	754.99	0	
754	Total n-3 Polyunsaturated (Omega-3) Fatty Acids	0148	0.3700	0.0000	0.3775	0.0171	0.0000	4	754.02	0	
754	Total n-3 Polyunsaturated (Omega-3) Fatty Acids	0676	0.3800	0.0000	0.3775	0.0171	0.0000	4	754.99	0	
754	Total n-3 Polyunsaturated (Omega-3) Fatty Acids	0610	0.4000	0.0000	0.3775	0.0171	0.0000	4	754.99	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Acids	0164	1.865	0.0300				3	756.99	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Acids	0148	1.895	0.0300				3	756.01	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Acids	0676	1.955	0.0100				3	756.99	0	

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
758	Total Saturated Fatty Acids (% (w/w))	0164	0.7600	0.0000				1	758.99	0	
762	Total Monounsaturated Fatty Acids (% (w/w))	0164	1.130	0.0000				1	762.99	0	
766	Total Polyunsaturated Fatty Acids (% (w/w))	0164	2.230	0.0400				1	766.99	0	
770	Total Fat (equivalent to NLEA) (% (w/w))	0164	4.325	0.0500				1	770.99	0	
772	Total Fatty Acids (% (w/w))	0226	4.035	0.1300				2	772.99	0	
772	Total Fatty Acids (% (w/w))	0164	4.140	0.0400				2	772.99	0	

Notes: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2 and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = Rejected for duplicates too far apart, 2 = Rejected as outlier, 5 = A reporting limit and 4 = zeros submitted as values. Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). Flag 3 indicates not used in statistics.