

**Animal Feed Scheme**

**Goat Feed, Medicated**

**Test Material Code # 202025**

**# Labs Reporting: 157**

**# Analytes Reported 114**

**Issue Date : 06/30/2020**

**Analyte Proficiency Testing Report**

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
000	Urea (%)	0278	0.9000	0.0000				1		000.02	0
000	Urea (%)	2302	0.0000	0.0000				1		000.02	4
001	Loss on Drying (%)	0278	7.420	0.0400	9.734	0.3681	0.1332	68	-6.28	001.07	0
001	Loss on Drying (%)	0618	7.590	0.2600	9.734	0.3681	0.1332	68	-5.82	001.99	0
001	Loss on Drying (%)	2155	8.490	0.1800	9.734	0.3681	0.1332	68	-3.38	001.99	0
001	Loss on Drying (%)	0510	8.650	0.1000	9.734	0.3681	0.1332	68	-2.94	001.99	0
001	Loss on Drying (%)	0019	8.850	0.3000	9.734	0.3681	0.1332	68	-2.40	001.07	0
001	Loss on Drying (%)	0038	8.910	0.2400	9.734	0.3681	0.1332	68	-2.24	001.07	0
001	Loss on Drying (%)	2081	9.200	0.1000	9.734	0.3681	0.1332	68	-1.45	001.99	0
001	Loss on Drying (%)	0536	9.225	0.0300	9.734	0.3681	0.1332	68	-1.38	001.99	0
001	Loss on Drying (%)	0353	9.250	0.2600	9.734	0.3681	0.1332	68	-1.31	001.07	0
001	Loss on Drying (%)	0853	9.250	0.1000	9.734	0.3681	0.1332	68	-1.31	001.07	0
001	Loss on Drying (%)	0889	9.285	0.0100	9.734	0.3681	0.1332	68	-1.22	001.99	0
001	Loss on Drying (%)	0169	9.310	0.0200	9.734	0.3681	0.1332	68	-1.15	001.00	0
001	Loss on Drying (%)	0034	9.345	0.2100	9.734	0.3681	0.1332	68	-1.06	001.07	0
001	Loss on Drying (%)	0089	9.355	0.0300	9.734	0.3681	0.1332	68	-1.03	001.07	0
001	Loss on Drying (%)	0918	9.440	0.1600	9.734	0.3681	0.1332	68	-0.80	001.99	0
001	Loss on Drying (%)	0083	9.450	0.0200	9.734	0.3681	0.1332	68	-0.77	001.07	0
001	Loss on Drying (%)	0676	9.465	0.9900	9.734	0.3681	0.1332	68	-0.73	001.99	0
001	Loss on Drying (%)	0948	9.490	0.0000	9.734	0.3681	0.1332	68	-0.66	001.99	0
001	Loss on Drying (%)	0345	9.530	0.0400	9.734	0.3681	0.1332	68	-0.55	001.07	0
001	Loss on Drying (%)	2146	9.565	0.1500	9.734	0.3681	0.1332	68	-0.46	001.07	0
001	Loss on Drying (%)	0142	9.570	0.0800	9.734	0.3681	0.1332	68	-0.44	001.07	0
001	Loss on Drying (%)	0675	9.575	0.0100	9.734	0.3681	0.1332	68	-0.43	001.07	0
001	Loss on Drying (%)	0868	9.600	0.0200	9.734	0.3681	0.1332	68	-0.36	001.03	0
001	Loss on Drying (%)	0171	9.605	0.0100	9.734	0.3681	0.1332	68	-0.35	001.07	0
001	Loss on Drying (%)	0186	9.614	0.0140	9.734	0.3681	0.1332	68	-0.33	001.99	0
001	Loss on Drying (%)	2259	9.650	0.0470	9.734	0.3681	0.1332	68	-0.23	001.07	0
001	Loss on Drying (%)	2144	9.670	0.0600	9.734	0.3681	0.1332	68	-0.17	001.07	0
001	Loss on Drying (%)	0226	9.680	0.2800	9.734	0.3681	0.1332	68	-0.15	001.07	0
001	Loss on Drying (%)	2192	9.715	0.0300	9.734	0.3681	0.1332	68	-0.05	001.07	0
001	Loss on Drying (%)	0682	9.730	0.0000	9.734	0.3681	0.1332	68	-0.01	001.07	0
001	Loss on Drying (%)	2295	9.815	0.0100	9.734	0.3681	0.1332	68	0.22	001.03	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
001	Loss on Drying (%)	0035	9.830	0.0400	9.734	0.3681	0.1332	68	0.26	001.07	0
001	Loss on Drying (%)	0612	9.840	0.2000	9.734	0.3681	0.1332	68	0.29	001.99	0
001	Loss on Drying (%)	0656	9.845	0.1100	9.734	0.3681	0.1332	68	0.30	001.99	0
001	Loss on Drying (%)	2059	9.850	0.0200	9.734	0.3681	0.1332	68	0.32	001.07	0
001	Loss on Drying (%)	0910	9.850	0.1000	9.734	0.3681	0.1332	68	0.32	001.99	0
001	Loss on Drying (%)	0309	9.852	0.1252	9.734	0.3681	0.1332	68	0.32	001.00	0
001	Loss on Drying (%)	0638	9.852	0.0660	9.734	0.3681	0.1332	68	0.32	001.07	0
001	Loss on Drying (%)	0074	9.855	0.1900	9.734	0.3681	0.1332	68	0.33	001.07	0
001	Loss on Drying (%)	0683	9.855	0.0100	9.734	0.3681	0.1332	68	0.33	001.07	0
001	Loss on Drying (%)	0028	9.865	0.1100	9.734	0.3681	0.1332	68	0.36	001.07	0
001	Loss on Drying (%)	2053	9.865	0.0900	9.734	0.3681	0.1332	68	0.36	001.07	0
001	Loss on Drying (%)	2012	9.870	0.0000	9.734	0.3681	0.1332	68	0.37	001.07	0
001	Loss on Drying (%)	0297	9.870	0.6000	9.734	0.3681	0.1332	68	0.37	001.07	0
001	Loss on Drying (%)	0590	9.875	0.1700	9.734	0.3681	0.1332	68	0.38	001.07	0
001	Loss on Drying (%)	2109	9.875	0.1900	9.734	0.3681	0.1332	68	0.38	001.07	0
001	Loss on Drying (%)	0505	9.880	0.0600	9.734	0.3681	0.1332	68	0.40	001.07	0
001	Loss on Drying (%)	0407	9.903	0.1210	9.734	0.3681	0.1332	68	0.46	001.07	0
001	Loss on Drying (%)	0037	9.915	0.0100	9.734	0.3681	0.1332	68	0.49	001.99	0
001	Loss on Drying (%)	0876	9.950	0.1000	9.734	0.3681	0.1332	68	0.59	001.07	0
001	Loss on Drying (%)	0693	9.958	0.1760	9.734	0.3681	0.1332	68	0.61	001.07	0
001	Loss on Drying (%)	0139	9.985	0.0300	9.734	0.3681	0.1332	68	0.68	001.07	0
001	Loss on Drying (%)	2062	9.990	0.0119	9.734	0.3681	0.1332	68	0.70	001.03	0
001	Loss on Drying (%)	0878	9.995	0.1500	9.734	0.3681	0.1332	68	0.71	001.03	0
001	Loss on Drying (%)	0571	10.03	0.1600	9.734	0.3681	0.1332	68	0.80	001.07	0
001	Loss on Drying (%)	0357	10.03	0.1842	9.734	0.3681	0.1332	68	0.82	001.99	0
001	Loss on Drying (%)	0843	10.06	0.4200	9.734	0.3681	0.1332	68	0.89	001.03	0
001	Loss on Drying (%)	0843	10.06	0.4200	9.734	0.3681	0.1332	68	0.89	001.07	0
001	Loss on Drying (%)	0581	10.07	0.0700	9.734	0.3681	0.1332	68	0.90	001.07	0
001	Loss on Drying (%)	0045	10.12	0.5900	9.734	0.3681	0.1332	68	1.04	001.07	0
001	Loss on Drying (%)	0148	10.15	0.1000	9.734	0.3681	0.1332	68	1.13	001.07	0
001	Loss on Drying (%)	0560	10.15	0.1000	9.734	0.3681	0.1332	68	1.13	001.99	0
001	Loss on Drying (%)	0199	10.16	0.0100	9.734	0.3681	0.1332	68	1.14	001.07	0
001	Loss on Drying (%)	0098	10.16	0.0000	9.734	0.3681	0.1332	68	1.16	001.07	0
001	Loss on Drying (%)	0872	10.16	0.1200	9.734	0.3681	0.1332	68	1.16	001.07	0
001	Loss on Drying (%)	0019	10.32	0.0600	9.734	0.3681	0.1332	68	1.59	001.99	0
001	Loss on Drying (%)	2304	10.33	0.3000	9.734	0.3681	0.1332	68	1.62	001.99	0
001	Loss on Drying (%)	0968	10.58	0.0420	9.734	0.3681	0.1332	68	2.29	001.05	0
001	Loss on Drying (%)	0003	7.955	3.730	9.734	0.3681	0.1332	68	-4.83	001.07	1
002	Protein, Crude (%)	0889	15.37	0.2500	17.20	0.2757	0.1464	153	-6.63	002.11	0
002	Protein, Crude (%)	0910	16.10	0.0000	17.20	0.2757	0.1464	153	-3.98	002.01	0
002	Protein, Crude (%)	0553	16.37	0.1400	17.20	0.2757	0.1464	153	-3.00	002.11	0
002	Protein, Crude (%)	2303	16.53	0.1500	17.20	0.2757	0.1464	153	-2.44	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0876	16.55	0.3000	17.20	0.2757	0.1464	153	-2.35	002.06	0
002	Protein, Crude (%)	0598	16.62	0.0200	17.20	0.2757	0.1464	153	-2.09	002.06	0
002	Protein, Crude (%)	0676	16.63	0.1300	17.20	0.2757	0.1464	153	-2.07	002.06	0
002	Protein, Crude (%)	0199	16.66	0.0300	17.20	0.2757	0.1464	153	-1.97	002.01	0
002	Protein, Crude (%)	0529	16.66	0.1900	17.20	0.2757	0.1464	153	-1.97	002.06	0
002	Protein, Crude (%)	2006	16.68	0.1400	17.20	0.2757	0.1464	153	-1.87	002.05	0
002	Protein, Crude (%)	0939	16.73	0.6000	17.20	0.2757	0.1464	153	-1.69	002.05	0
002	Protein, Crude (%)	2291	16.74	0.0300	17.20	0.2757	0.1464	153	-1.68	002.05	0
002	Protein, Crude (%)	0169	16.76	0.3800	17.20	0.2757	0.1464	153	-1.58	002.00	0
002	Protein, Crude (%)	0910	16.80	0.2000	17.20	0.2757	0.1464	153	-1.44	002.06	0
002	Protein, Crude (%)	0144	16.83	0.1700	17.20	0.2757	0.1464	153	-1.35	002.06	0
002	Protein, Crude (%)	0646	16.86	0.1300	17.20	0.2757	0.1464	153	-1.24	002.06	0
002	Protein, Crude (%)	0011	16.88	0.0000	17.20	0.2757	0.1464	153	-1.15	002.11	0
002	Protein, Crude (%)	0589	16.89	0.1000	17.20	0.2757	0.1464	153	-1.11	002.06	0
002	Protein, Crude (%)	0100	16.90	0.0700	17.20	0.2757	0.1464	153	-1.09	002.06	0
002	Protein, Crude (%)	0171	16.91	0.0400	17.20	0.2757	0.1464	153	-1.04	002.06	0
002	Protein, Crude (%)	0074	16.92	0.1500	17.20	0.2757	0.1464	153	-1.02	002.06	0
002	Protein, Crude (%)	0148	16.93	0.0440	17.20	0.2757	0.1464	153	-0.97	002.06	0
002	Protein, Crude (%)	0553	16.94	0.6000	17.20	0.2757	0.1464	153	-0.93	002.06	0
002	Protein, Crude (%)	0536	16.95	0.0800	17.20	0.2757	0.1464	153	-0.90	002.05	0
002	Protein, Crude (%)	0036	16.97	0.0728	17.20	0.2757	0.1464	153	-0.84	002.02	0
002	Protein, Crude (%)	0870	16.97	0.0630	17.20	0.2757	0.1464	153	-0.82	002.01	0
002	Protein, Crude (%)	0066	16.98	0.0900	17.20	0.2757	0.1464	153	-0.80	002.02	0
002	Protein, Crude (%)	0297	16.98	0.0200	17.20	0.2757	0.1464	153	-0.79	002.11	0
002	Protein, Crude (%)	0885	16.98	0.0409	17.20	0.2757	0.1464	153	-0.77	002.05	0
002	Protein, Crude (%)	0953	16.99	0.2500	17.20	0.2757	0.1464	153	-0.77	002.06	0
002	Protein, Crude (%)	0123	16.99	0.0000	17.20	0.2757	0.1464	153	-0.75	002.06	0
002	Protein, Crude (%)	2109	16.99	0.0200	17.20	0.2757	0.1464	153	-0.75	002.05	0
002	Protein, Crude (%)	0510	17.00	0.0000	17.20	0.2757	0.1464	153	-0.71	002.06	0
002	Protein, Crude (%)	0148	17.00	0.0000	17.20	0.2757	0.1464	153	-0.71	002.11	0
002	Protein, Crude (%)	2295	17.01	0.0100	17.20	0.2757	0.1464	153	-0.70	002.05	0
002	Protein, Crude (%)	0881	17.01	0.1040	17.20	0.2757	0.1464	153	-0.68	002.05	0
002	Protein, Crude (%)	0226	17.02	0.0600	17.20	0.2757	0.1464	153	-0.64	002.06	0
002	Protein, Crude (%)	0354	17.03	0.0300	17.20	0.2757	0.1464	153	-0.62	002.05	0
002	Protein, Crude (%)	2302	17.03	0.1900	17.20	0.2757	0.1464	153	-0.62	002.11	0
002	Protein, Crude (%)	0505	17.03	0.2600	17.20	0.2757	0.1464	153	-0.61	002.06	0
002	Protein, Crude (%)	0358	17.04	0.6400	17.20	0.2757	0.1464	153	-0.57	002.06	0
002	Protein, Crude (%)	2062	17.04	0.0680	17.20	0.2757	0.1464	153	-0.56	002.05	0
002	Protein, Crude (%)	0878	17.05	0.0700	17.20	0.2757	0.1464	153	-0.55	002.05	0
002	Protein, Crude (%)	0868	17.05	0.0600	17.20	0.2757	0.1464	153	-0.53	002.05	0
002	Protein, Crude (%)	2053	17.05	0.3000	17.20	0.2757	0.1464	153	-0.53	002.06	0
002	Protein, Crude (%)	0683	17.05	0.0400	17.20	0.2757	0.1464	153	-0.53	002.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0278	17.05	0.1000	17.20	0.2757	0.1464	153	-0.53	002.06	0
002	Protein, Crude (%)	2192	17.06	0.0400	17.20	0.2757	0.1464	153	-0.50	002.05	0
002	Protein, Crude (%)	2188	17.07	0.0500	17.20	0.2757	0.1464	153	-0.48	002.01	0
002	Protein, Crude (%)	0948	17.07	0.0400	17.20	0.2757	0.1464	153	-0.46	002.05	0
002	Protein, Crude (%)	2150	17.08	0.1200	17.20	0.2757	0.1464	153	-0.42	002.06	0
002	Protein, Crude (%)	2012	17.10	0.1100	17.20	0.2757	0.1464	153	-0.37	002.05	0
002	Protein, Crude (%)	0652	17.10	0.0000	17.20	0.2757	0.1464	153	-0.35	002.01	0
002	Protein, Crude (%)	0728	17.10	0.1140	17.20	0.2757	0.1464	153	-0.35	002.04	0
002	Protein, Crude (%)	0014	17.10	0.2000	17.20	0.2757	0.1464	153	-0.35	002.06	0
002	Protein, Crude (%)	0884	17.10	0.0000	17.20	0.2757	0.1464	153	-0.35	002.06	0
002	Protein, Crude (%)	0643	17.11	0.2100	17.20	0.2757	0.1464	153	-0.33	002.99	0
002	Protein, Crude (%)	0870	17.11	0.1069	17.20	0.2757	0.1464	153	-0.33	002.06	0
002	Protein, Crude (%)	0038	17.11	0.4640	17.20	0.2757	0.1464	153	-0.32	002.06	0
002	Protein, Crude (%)	2146	17.11	0.1200	17.20	0.2757	0.1464	153	-0.32	002.05	0
002	Protein, Crude (%)	0037	17.11	0.1400	17.20	0.2757	0.1464	153	-0.32	002.06	0
002	Protein, Crude (%)	0199	17.11	0.1600	17.20	0.2757	0.1464	153	-0.32	002.06	0
002	Protein, Crude (%)	0028	17.12	0.1500	17.20	0.2757	0.1464	153	-0.30	002.06	0
002	Protein, Crude (%)	0202	17.12	0.1300	17.20	0.2757	0.1464	153	-0.30	002.06	0
002	Protein, Crude (%)	0848	17.13	0.0800	17.20	0.2757	0.1464	153	-0.24	002.06	0
002	Protein, Crude (%)	0638	17.13	0.0070	17.20	0.2757	0.1464	153	-0.23	002.04	0
002	Protein, Crude (%)	0353	17.15	0.1700	17.20	0.2757	0.1464	153	-0.19	002.06	0
002	Protein, Crude (%)	0968	17.15	0.0860	17.20	0.2757	0.1464	153	-0.17	002.01	0
002	Protein, Crude (%)	0003	17.15	0.0000	17.20	0.2757	0.1464	153	-0.17	002.06	0
002	Protein, Crude (%)	0010	17.15	0.5000	17.20	0.2757	0.1464	153	-0.17	002.06	0
002	Protein, Crude (%)	0139	17.15	0.0800	17.20	0.2757	0.1464	153	-0.17	002.06	0
002	Protein, Crude (%)	0175	17.15	0.1000	17.20	0.2757	0.1464	153	-0.17	002.06	0
002	Protein, Crude (%)	0939	17.15	0.0400	17.20	0.2757	0.1464	153	-0.17	002.06	0
002	Protein, Crude (%)	0675	17.17	0.0700	17.20	0.2757	0.1464	153	-0.12	002.05	0
002	Protein, Crude (%)	2073	17.17	0.3200	17.20	0.2757	0.1464	153	-0.10	002.05	0
002	Protein, Crude (%)	0089	17.17	0.0000	17.20	0.2757	0.1464	153	-0.10	002.06	0
002	Protein, Crude (%)	0298	17.17	0.0200	17.20	0.2757	0.1464	153	-0.10	002.06	0
002	Protein, Crude (%)	0948	17.17	0.0000	17.20	0.2757	0.1464	153	-0.10	002.06	0
002	Protein, Crude (%)	0297	17.18	0.0400	17.20	0.2757	0.1464	153	-0.06	002.06	0
002	Protein, Crude (%)	0574	17.18	0.0600	17.20	0.2757	0.1464	153	-0.06	002.06	0
002	Protein, Crude (%)	2109	17.19	0.2100	17.20	0.2757	0.1464	153	-0.04	002.06	0
002	Protein, Crude (%)	0354	17.19	0.0100	17.20	0.2757	0.1464	153	-0.04	002.06	0
002	Protein, Crude (%)	0693	17.19	0.2990	17.20	0.2757	0.1464	153	-0.03	002.06	0
002	Protein, Crude (%)	0425	17.19	0.0000	17.20	0.2757	0.1464	153	-0.02	002.06	0
002	Protein, Crude (%)	0009	17.20	0.3800	17.20	0.2757	0.1464	153	0.01	002.06	0
002	Protein, Crude (%)	2168	17.21	0.0600	17.20	0.2757	0.1464	153	0.05	002.06	0
002	Protein, Crude (%)	0263	17.21	0.0280	17.20	0.2757	0.1464	153	0.06	002.06	0
002	Protein, Crude (%)	0941	17.22	0.1900	17.20	0.2757	0.1464	153	0.07	002.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0001	17.22	0.3300	17.20	0.2757	0.1464	153	0.07	002.06	0
002	Protein, Crude (%)	0142	17.23	0.0100	17.20	0.2757	0.1464	153	0.10	002.06	0
002	Protein, Crude (%)	0626	17.23	0.0700	17.20	0.2757	0.1464	153	0.10	002.06	0
002	Protein, Crude (%)	2196	17.23	0.0000	17.20	0.2757	0.1464	153	0.12	002.01	0
002	Protein, Crude (%)	0098	17.25	0.0300	17.20	0.2757	0.1464	153	0.17	002.08	0
002	Protein, Crude (%)	0034	17.25	0.1780	17.20	0.2757	0.1464	153	0.18	002.06	0
002	Protein, Crude (%)	0229	17.25	0.3200	17.20	0.2757	0.1464	153	0.19	002.06	0
002	Protein, Crude (%)	0590	17.25	0.3000	17.20	0.2757	0.1464	153	0.19	002.06	0
002	Protein, Crude (%)	0857	17.27	0.4300	17.20	0.2757	0.1464	153	0.25	002.06	0
002	Protein, Crude (%)	2181	17.27	0.1000	17.20	0.2757	0.1464	153	0.27	002.05	0
002	Protein, Crude (%)	0682	17.28	0.0000	17.20	0.2757	0.1464	153	0.30	002.06	0
002	Protein, Crude (%)	2144	17.30	0.3000	17.20	0.2757	0.1464	153	0.37	002.05	0
002	Protein, Crude (%)	0674	17.30	0.0000	17.20	0.2757	0.1464	153	0.37	002.05	0
002	Protein, Crude (%)	0098	17.30	0.0000	17.20	0.2757	0.1464	153	0.37	002.06	0
002	Protein, Crude (%)	2174	17.31	0.1300	17.20	0.2757	0.1464	153	0.39	002.06	0
002	Protein, Crude (%)	0872	17.31	0.1750	17.20	0.2757	0.1464	153	0.42	002.06	0
002	Protein, Crude (%)	0914	17.32	0.0700	17.20	0.2757	0.1464	153	0.43	002.06	0
002	Protein, Crude (%)	0889	17.33	0.2000	17.20	0.2757	0.1464	153	0.48	002.01	0
002	Protein, Crude (%)	0407	17.33	0.1980	17.20	0.2757	0.1464	153	0.49	002.06	0
002	Protein, Crude (%)	0309	17.33	0.0870	17.20	0.2757	0.1464	153	0.50	002.06	0
002	Protein, Crude (%)	0571	17.34	0.1500	17.20	0.2757	0.1464	153	0.50	002.06	0
002	Protein, Crude (%)	0035	17.36	0.1960	17.20	0.2757	0.1464	153	0.60	002.06	0
002	Protein, Crude (%)	0019	17.37	0.3400	17.20	0.2757	0.1464	153	0.63	002.06	0
002	Protein, Crude (%)	0868	17.37	0.0400	17.20	0.2757	0.1464	153	0.63	002.06	0
002	Protein, Crude (%)	2059	17.37	0.0718	17.20	0.2757	0.1464	153	0.64	002.06	0
002	Protein, Crude (%)	0357	17.38	0.1250	17.20	0.2757	0.1464	153	0.65	002.06	0
002	Protein, Crude (%)	0726	17.38	0.0800	17.20	0.2757	0.1464	153	0.66	002.06	0
002	Protein, Crude (%)	0956	17.40	0.2000	17.20	0.2757	0.1464	153	0.74	002.06	0
002	Protein, Crude (%)	2184	17.41	0.2300	17.20	0.2757	0.1464	153	0.76	002.06	0
002	Protein, Crude (%)	2033	17.42	0.1400	17.20	0.2757	0.1464	153	0.81	002.06	0
002	Protein, Crude (%)	0541	17.43	0.0300	17.20	0.2757	0.1464	153	0.83	002.06	0
002	Protein, Crude (%)	2012	17.43	0.3600	17.20	0.2757	0.1464	153	0.85	002.06	0
002	Protein, Crude (%)	0083	17.44	0.1100	17.20	0.2757	0.1464	153	0.86	002.06	0
002	Protein, Crude (%)	0045	17.45	0.1000	17.20	0.2757	0.1464	153	0.92	002.06	0
002	Protein, Crude (%)	0294	17.45	0.1000	17.20	0.2757	0.1464	153	0.92	002.06	0
002	Protein, Crude (%)	0042	17.45	0.3000	17.20	0.2757	0.1464	153	0.92	002.06	0
002	Protein, Crude (%)	0208	17.45	0.3000	17.20	0.2757	0.1464	153	0.92	002.06	0
002	Protein, Crude (%)	0918	17.46	0.0400	17.20	0.2757	0.1464	153	0.95	002.06	0
002	Protein, Crude (%)	0018	17.48	0.2700	17.20	0.2757	0.1464	153	1.01	002.06	0
002	Protein, Crude (%)	0520	17.50	0.2000	17.20	0.2757	0.1464	153	1.10	002.06	0
002	Protein, Crude (%)	0687	17.50	0.4000	17.20	0.2757	0.1464	153	1.10	002.06	0
002	Protein, Crude (%)	0692	17.50	0.6000	17.20	0.2757	0.1464	153	1.10	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0968	17.50	0.1130	17.20	0.2757	0.1464	153	1.11	002.06	0
002	Protein, Crude (%)	0036	17.51	0.2900	17.20	0.2757	0.1464	153	1.12	002.06	0
002	Protein, Crude (%)	0675	17.51	0.0900	17.20	0.2757	0.1464	153	1.12	002.06	0
002	Protein, Crude (%)	0345	17.53	0.1500	17.20	0.2757	0.1464	153	1.19	002.06	0
002	Protein, Crude (%)	0581	17.53	0.5700	17.20	0.2757	0.1464	153	1.19	002.06	0
002	Protein, Crude (%)	2129	17.54	0.0670	17.20	0.2757	0.1464	153	1.26	002.06	0
002	Protein, Crude (%)	0265	17.55	0.3000	17.20	0.2757	0.1464	153	1.28	002.06	0
002	Protein, Crude (%)	2178	17.55	0.4000	17.20	0.2757	0.1464	153	1.28	002.06	0
002	Protein, Crude (%)	2076	17.56	0.0880	17.20	0.2757	0.1464	153	1.31	002.06	0
002	Protein, Crude (%)	0859	17.57	0.0900	17.20	0.2757	0.1464	153	1.36	002.06	0
002	Protein, Crude (%)	0190	17.62	0.0100	17.20	0.2757	0.1464	153	1.52	002.06	0
002	Protein, Crude (%)	2081	17.64	0.1200	17.20	0.2757	0.1464	153	1.61	002.06	0
002	Protein, Crude (%)	0164	17.66	0.0600	17.20	0.2757	0.1464	153	1.68	002.06	0
002	Protein, Crude (%)	0843	17.68	0.1000	17.20	0.2757	0.1464	153	1.75	002.06	0
002	Protein, Crude (%)	0186	17.70	0.3700	17.20	0.2757	0.1464	153	1.81	002.06	0
002	Protein, Crude (%)	2302	17.77	0.2800	17.20	0.2757	0.1464	153	2.08	002.06	0
002	Protein, Crude (%)	0618	17.79	0.3050	17.20	0.2757	0.1464	153	2.14	002.06	0
002	Protein, Crude (%)	0656	17.80	0.0000	17.20	0.2757	0.1464	153	2.19	002.06	0
002	Protein, Crude (%)	2089	17.90	0.0100	17.20	0.2757	0.1464	153	2.53	002.06	0
002	Protein, Crude (%)	2155	18.23	0.0950	17.20	0.2757	0.1464	153	3.76	002.06	0
002	Protein, Crude (%)	0853	18.45	0.1000	17.20	0.2757	0.1464	153	4.55	002.06	0
002	Protein, Crude (%)	0612	19.39	0.1000	17.20	0.2757	0.1464	153	7.96	002.06	0
002	Protein, Crude (%)	0964	19.64	0.0440	17.20	0.2757	0.1464	153	8.86	002.06	0
002	Protein, Crude (%)	2304	16.97	0.8200	17.20	0.2757	0.1464	153	-0.82	002.06	1
002	Protein, Crude (%)	0921	13.28	0.1800	17.20	0.2757	0.1464	153	-14.21	002.06	2
003	Fat, Crude (%)	0144	3.835	0.1500	4.627	0.3253	0.0871	118	-2.44	003.14	0
003	Fat, Crude (%)	2291	3.930	0.0600	4.627	0.3253	0.0871	118	-2.14	003.13	0
003	Fat, Crude (%)	0357	3.950	0.1000	4.627	0.3253	0.0871	118	-2.08	003.14	0
003	Fat, Crude (%)	0885	3.998	0.0616	4.627	0.3253	0.0871	118	-1.93	003.00	0
003	Fat, Crude (%)	0298	4.000	0.0200	4.627	0.3253	0.0871	118	-1.93	003.14	0
003	Fat, Crude (%)	0175	4.020	0.2800	4.627	0.3253	0.0871	118	-1.87	003.14	0
003	Fat, Crude (%)	2304	4.025	0.1300	4.627	0.3253	0.0871	118	-1.85	003.99	0
003	Fat, Crude (%)	0035	4.090	0.0600	4.627	0.3253	0.0871	118	-1.65	003.14	0
003	Fat, Crude (%)	0876	4.100	0.0000	4.627	0.3253	0.0871	118	-1.62	003.14	0
003	Fat, Crude (%)	2076	4.120	0.2219	4.627	0.3253	0.0871	118	-1.56	003.14	0
003	Fat, Crude (%)	0843	4.130	0.0600	4.627	0.3253	0.0871	118	-1.53	003.14	0
003	Fat, Crude (%)	0868	4.135	0.0900	4.627	0.3253	0.0871	118	-1.51	003.14	0
003	Fat, Crude (%)	0939	4.145	0.0300	4.627	0.3253	0.0871	118	-1.48	003.06	0
003	Fat, Crude (%)	2302	4.190	0.4200	4.627	0.3253	0.0871	118	-1.34	003.14	0
003	Fat, Crude (%)	0036	4.223	0.2624	4.627	0.3253	0.0871	118	-1.24	003.10	0
003	Fat, Crude (%)	0083	4.230	0.0600	4.627	0.3253	0.0871	118	-1.22	003.14	0
003	Fat, Crude (%)	0536	4.255	0.0500	4.627	0.3253	0.0871	118	-1.14	003.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
003	Fat, Crude (%)	0505	4.260	0.0800	4.627	0.3253	0.0871	118	-1.13	003.14	0
003	Fat, Crude (%)	0148	4.285	0.1700	4.627	0.3253	0.0871	118	-1.05	003.14	0
003	Fat, Crude (%)	0425	4.300	0.0000	4.627	0.3253	0.0871	118	-1.01	003.06	0
003	Fat, Crude (%)	0853	4.300	0.0000	4.627	0.3253	0.0871	118	-1.01	003.14	0
003	Fat, Crude (%)	2073	4.315	0.0300	4.627	0.3253	0.0871	118	-0.96	003.10	0
003	Fat, Crude (%)	0529	4.350	0.3200	4.627	0.3253	0.0871	118	-0.85	003.14	0
003	Fat, Crude (%)	0956	4.350	0.1000	4.627	0.3253	0.0871	118	-0.85	003.14	0
003	Fat, Crude (%)	0581	4.360	0.1000	4.627	0.3253	0.0871	118	-0.82	003.14	0
003	Fat, Crude (%)	0089	4.390	0.0200	4.627	0.3253	0.0871	118	-0.73	003.14	0
003	Fat, Crude (%)	0171	4.395	0.0100	4.627	0.3253	0.0871	118	-0.71	003.12	0
003	Fat, Crude (%)	0202	4.395	0.0500	4.627	0.3253	0.0871	118	-0.71	003.14	0
003	Fat, Crude (%)	0918	4.415	0.0100	4.627	0.3253	0.0871	118	-0.65	003.06	0
003	Fat, Crude (%)	0574	4.420	0.0400	4.627	0.3253	0.0871	118	-0.64	003.14	0
003	Fat, Crude (%)	2062	4.420	0.0371	4.627	0.3253	0.0871	118	-0.64	003.12	0
003	Fat, Crude (%)	0265	4.440	0.0000	4.627	0.3253	0.0871	118	-0.58	003.14	0
003	Fat, Crude (%)	0407	4.462	0.0965	4.627	0.3253	0.0871	118	-0.51	003.14	0
003	Fat, Crude (%)	0941	4.475	0.2300	4.627	0.3253	0.0871	118	-0.47	003.99	0
003	Fat, Crude (%)	2192	4.495	0.0300	4.627	0.3253	0.0871	118	-0.41	003.00	0
003	Fat, Crude (%)	0229	4.495	0.3500	4.627	0.3253	0.0871	118	-0.41	003.14	0
003	Fat, Crude (%)	0148	4.500	0.0000	4.627	0.3253	0.0871	118	-0.39	003.11	0
003	Fat, Crude (%)	0186	4.500	0.0200	4.627	0.3253	0.0871	118	-0.39	003.14	0
003	Fat, Crude (%)	0693	4.511	0.2160	4.627	0.3253	0.0871	118	-0.36	003.10	0
003	Fat, Crude (%)	0062	4.513	0.0150	4.627	0.3253	0.0871	118	-0.35	003.10	0
003	Fat, Crude (%)	0263	4.514	0.0340	4.627	0.3253	0.0871	118	-0.35	003.10	0
003	Fat, Crude (%)	0294	4.525	0.0900	4.627	0.3253	0.0871	118	-0.31	003.06	0
003	Fat, Crude (%)	0638	4.526	0.1490	4.627	0.3253	0.0871	118	-0.31	003.09	0
003	Fat, Crude (%)	0553	4.535	0.0500	4.627	0.3253	0.0871	118	-0.28	003.11	0
003	Fat, Crude (%)	2188	4.555	0.0900	4.627	0.3253	0.0871	118	-0.22	003.10	0
003	Fat, Crude (%)	2089	4.555	0.1100	4.627	0.3253	0.0871	118	-0.22	003.13	0
003	Fat, Crude (%)	0019	4.565	0.1300	4.627	0.3253	0.0871	118	-0.19	003.14	0
003	Fat, Crude (%)	2184	4.575	0.0500	4.627	0.3253	0.0871	118	-0.16	003.10	0
003	Fat, Crude (%)	2302	4.585	0.0300	4.627	0.3253	0.0871	118	-0.13	003.11	0
003	Fat, Crude (%)	0541	4.590	0.0000	4.627	0.3253	0.0871	118	-0.11	003.10	0
003	Fat, Crude (%)	0728	4.592	0.0240	4.627	0.3253	0.0871	118	-0.11	003.10	0
003	Fat, Crude (%)	0646	4.595	0.0900	4.627	0.3253	0.0871	118	-0.10	003.13	0
003	Fat, Crude (%)	0358	4.595	0.0300	4.627	0.3253	0.0871	118	-0.10	003.14	0
003	Fat, Crude (%)	0870	4.599	0.0902	4.627	0.3253	0.0871	118	-0.09	003.10	0
003	Fat, Crude (%)	0345	4.635	0.1300	4.627	0.3253	0.0871	118	0.02	003.00	0
003	Fat, Crude (%)	0297	4.635	0.1300	4.627	0.3253	0.0871	118	0.02	003.06	0
003	Fat, Crude (%)	0045	4.635	0.0500	4.627	0.3253	0.0871	118	0.02	003.10	0
003	Fat, Crude (%)	2196	4.640	0.0000	4.627	0.3253	0.0871	118	0.04	003.10	0
003	Fat, Crude (%)	0878	4.645	0.0900	4.627	0.3253	0.0871	118	0.06	003.14	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
003	Fat, Crude (%)	0948	4.645	0.0500	4.627	0.3253	0.0871	118	0.06	003.14	0
003	Fat, Crude (%)	0278	4.650	0.1000	4.627	0.3253	0.0871	118	0.07	003.14	0
003	Fat, Crude (%)	2033	4.655	0.0300	4.627	0.3253	0.0871	118	0.09	003.06	0
003	Fat, Crude (%)	0098	4.665	0.0500	4.627	0.3253	0.0871	118	0.12	003.13	0
003	Fat, Crude (%)	0968	4.666	0.0020	4.627	0.3253	0.0871	118	0.12	003.14	0
003	Fat, Crude (%)	0948	4.670	0.0000	4.627	0.3253	0.0871	118	0.13	003.09	0
003	Fat, Crude (%)	0035	4.680	0.2200	4.627	0.3253	0.0871	118	0.16	003.00	0
003	Fat, Crude (%)	0683	4.680	0.1600	4.627	0.3253	0.0871	118	0.16	003.06	0
003	Fat, Crude (%)	0208	4.680	0.1200	4.627	0.3253	0.0871	118	0.16	003.13	0
003	Fat, Crude (%)	0889	4.698	0.0050	4.627	0.3253	0.0871	118	0.22	003.11	0
003	Fat, Crude (%)	0001	4.703	0.0890	4.627	0.3253	0.0871	118	0.23	003.14	0
003	Fat, Crude (%)	0028	4.705	0.0100	4.627	0.3253	0.0871	118	0.24	003.13	0
003	Fat, Crude (%)	2006	4.705	0.0900	4.627	0.3253	0.0871	118	0.24	003.13	0
003	Fat, Crude (%)	0074	4.720	0.2800	4.627	0.3253	0.0871	118	0.29	003.14	0
003	Fat, Crude (%)	2012	4.730	0.0200	4.627	0.3253	0.0871	118	0.32	003.09	0
003	Fat, Crude (%)	0066	4.735	0.1500	4.627	0.3253	0.0871	118	0.33	003.14	0
003	Fat, Crude (%)	2295	4.735	0.0100	4.627	0.3253	0.0871	118	0.33	003.06	0
003	Fat, Crude (%)	0037	4.760	0.0200	4.627	0.3253	0.0871	118	0.41	003.14	0
003	Fat, Crude (%)	0910	4.770	0.0200	4.627	0.3253	0.0871	118	0.44	003.06	0
003	Fat, Crude (%)	0598	4.780	0.0200	4.627	0.3253	0.0871	118	0.47	003.14	0
003	Fat, Crude (%)	0098	4.785	0.1100	4.627	0.3253	0.0871	118	0.49	003.10	0
003	Fat, Crude (%)	0309	4.786	0.0325	4.627	0.3253	0.0871	118	0.49	003.00	0
003	Fat, Crude (%)	0175	4.790	0.1000	4.627	0.3253	0.0871	118	0.50	003.00	0
003	Fat, Crude (%)	0510	4.800	0.0000	4.627	0.3253	0.0871	118	0.53	003.09	0
003	Fat, Crude (%)	0682	4.800	0.0000	4.627	0.3253	0.0871	118	0.53	003.14	0
003	Fat, Crude (%)	0100	4.820	0.0800	4.627	0.3253	0.0871	118	0.59	003.10	0
003	Fat, Crude (%)	0148	4.835	0.0100	4.627	0.3253	0.0871	118	0.64	003.10	0
003	Fat, Crude (%)	0726	4.835	0.0700	4.627	0.3253	0.0871	118	0.64	003.14	0
003	Fat, Crude (%)	0353	4.840	0.0800	4.627	0.3253	0.0871	118	0.65	003.09	0
003	Fat, Crude (%)	0354	4.850	0.0000	4.627	0.3253	0.0871	118	0.69	003.09	0
003	Fat, Crude (%)	0881	4.860	0.1490	4.627	0.3253	0.0871	118	0.71	003.10	0
003	Fat, Crude (%)	0164	4.860	0.1400	4.627	0.3253	0.0871	118	0.72	003.01	0
003	Fat, Crude (%)	0123	4.860	0.0000	4.627	0.3253	0.0871	118	0.72	003.13	0
003	Fat, Crude (%)	0674	4.870	0.0000	4.627	0.3253	0.0871	118	0.75	003.09	0
003	Fat, Crude (%)	0682	4.890	0.0000	4.627	0.3253	0.0871	118	0.81	003.06	0
003	Fat, Crude (%)	0199	4.895	0.0300	4.627	0.3253	0.0871	118	0.82	003.06	0
003	Fat, Crude (%)	0098	4.895	0.0100	4.627	0.3253	0.0871	118	0.82	003.09	0
003	Fat, Crude (%)	0687	4.900	0.2000	4.627	0.3253	0.0871	118	0.84	003.06	0
003	Fat, Crude (%)	0003	4.905	0.2500	4.627	0.3253	0.0871	118	0.85	003.14	0
003	Fat, Crude (%)	0889	4.915	0.2300	4.627	0.3253	0.0871	118	0.89	003.10	0
003	Fat, Crude (%)	2181	4.920	0.0600	4.627	0.3253	0.0871	118	0.90	003.06	0
003	Fat, Crude (%)	0870	4.924	0.0599	4.627	0.3253	0.0871	118	0.91	003.09	0



Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
003	Fat, Crude (%)	0169	4.955	0.0500	4.627	0.3253	0.0871	118	1.01	003.06	0
003	Fat, Crude (%)	0226	4.955	0.1500	4.627	0.3253	0.0871	118	1.01	003.09	0
003	Fat, Crude (%)	0229	4.960	0.0800	4.627	0.3253	0.0871	118	1.02	003.06	0
003	Fat, Crude (%)	0009	4.985	0.0500	4.627	0.3253	0.0871	118	1.10	003.14	0
003	Fat, Crude (%)	0590	5.000	0.0200	4.627	0.3253	0.0871	118	1.15	003.09	0
003	Fat, Crude (%)	2109	5.025	0.1900	4.627	0.3253	0.0871	118	1.22	003.00	0
003	Fat, Crude (%)	0675	5.025	0.0300	4.627	0.3253	0.0871	118	1.22	003.09	0
003	Fat, Crude (%)	2053	5.035	0.0900	4.627	0.3253	0.0871	118	1.25	003.00	0
003	Fat, Crude (%)	0618	5.182	0.6060	4.627	0.3253	0.0871	118	1.71	003.10	0
003	Fat, Crude (%)	0870	5.189	0.0974	4.627	0.3253	0.0871	118	1.73	003.13	0
003	Fat, Crude (%)	0656	5.190	0.0000	4.627	0.3253	0.0871	118	1.73	003.99	0
003	Fat, Crude (%)	0190	5.200	0.0400	4.627	0.3253	0.0871	118	1.76	003.14	0
003	Fat, Crude (%)	0011	5.220	0.0000	4.627	0.3253	0.0871	118	1.82	003.11	0
003	Fat, Crude (%)	0884	5.350	0.3000	4.627	0.3253	0.0871	118	2.22	003.06	0
003	Fat, Crude (%)	0297	5.495	0.0700	4.627	0.3253	0.0871	118	2.67	003.11	0
003	Fat, Crude (%)	0032	5.550	0.1000	4.627	0.3253	0.0871	118	2.84	003.00	0
003	Fat, Crude (%)	0520	5.650	0.0000	4.627	0.3253	0.0871	118	3.14	003.14	0
003	Fat, Crude (%)	0034	3.595	0.9100	4.627	0.3253	0.0871	118	-3.17	003.14	1
003	Fat, Crude (%)	2155	3.989	0.9180	4.627	0.3253	0.0871	118	-1.96	003.10	1
003	Fat, Crude (%)	0692	4.600	1.000	4.627	0.3253	0.0871	118	-0.08	003.14	1
003	Fat, Crude (%)	2129	1.409	0.0090	4.627	0.3253	0.0871	118	-9.90	003.14	2
004	Fiber, Crude (%)	0425	3.100	0.0000	5.875	0.3893	0.1431	101	-7.13	004.00	0
004	Fiber, Crude (%)	0202	4.535	0.0100	5.875	0.3893	0.1431	101	-3.44	004.07	0
004	Fiber, Crude (%)	0675	4.890	0.0600	5.875	0.3893	0.1431	101	-2.53	004.07	0
004	Fiber, Crude (%)	0656	5.030	0.0000	5.875	0.3893	0.1431	101	-2.17	004.07	0
004	Fiber, Crude (%)	0553	5.065	0.0500	5.875	0.3893	0.1431	101	-2.08	004.11	0
004	Fiber, Crude (%)	2146	5.090	0.1200	5.875	0.3893	0.1431	101	-2.02	004.00	0
004	Fiber, Crude (%)	2109	5.145	0.2100	5.875	0.3893	0.1431	101	-1.88	004.07	0
004	Fiber, Crude (%)	0265	5.295	0.0100	5.875	0.3893	0.1431	101	-1.49	004.07	0
004	Fiber, Crude (%)	0028	5.300	0.4000	5.875	0.3893	0.1431	101	-1.48	004.07	0
004	Fiber, Crude (%)	0510	5.350	0.1000	5.875	0.3893	0.1431	101	-1.35	004.07	0
004	Fiber, Crude (%)	0590	5.350	0.4400	5.875	0.3893	0.1431	101	-1.35	004.07	0
004	Fiber, Crude (%)	2302	5.390	0.0600	5.875	0.3893	0.1431	101	-1.25	004.11	0
004	Fiber, Crude (%)	0682	5.400	0.0000	5.875	0.3893	0.1431	101	-1.22	004.07	0
004	Fiber, Crude (%)	2089	5.435	0.0500	5.875	0.3893	0.1431	101	-1.13	004.03	0
004	Fiber, Crude (%)	0683	5.495	0.0900	5.875	0.3893	0.1431	101	-0.98	004.07	0
004	Fiber, Crude (%)	0968	5.530	0.1250	5.875	0.3893	0.1431	101	-0.89	004.07	0
004	Fiber, Crude (%)	0123	5.540	0.0400	5.875	0.3893	0.1431	101	-0.86	004.07	0
004	Fiber, Crude (%)	0175	5.545	0.0500	5.875	0.3893	0.1431	101	-0.85	004.00	0
004	Fiber, Crude (%)	0035	5.545	0.0100	5.875	0.3893	0.1431	101	-0.85	004.07	0
004	Fiber, Crude (%)	0726	5.545	0.0100	5.875	0.3893	0.1431	101	-0.85	004.07	0
004	Fiber, Crude (%)	0354	5.550	0.0600	5.875	0.3893	0.1431	101	-0.84	004.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber, Crude (%)	0190	5.550	0.1000	5.875	0.3893	0.1431	101	-0.84	004.07	0
004	Fiber, Crude (%)	0956	5.550	0.1000	5.875	0.3893	0.1431	101	-0.84	004.07	0
004	Fiber, Crude (%)	2053	5.565	0.0500	5.875	0.3893	0.1431	101	-0.80	004.07	0
004	Fiber, Crude (%)	0910	5.650	0.5000	5.875	0.3893	0.1431	101	-0.58	004.07	0
004	Fiber, Crude (%)	2304	5.660	0.3200	5.875	0.3893	0.1431	101	-0.55	004.07	0
004	Fiber, Crude (%)	0011	5.660	0.0000	5.875	0.3893	0.1431	101	-0.55	004.11	0
004	Fiber, Crude (%)	0407	5.665	0.4162	5.875	0.3893	0.1431	101	-0.54	004.07	0
004	Fiber, Crude (%)	2012	5.675	0.2900	5.875	0.3893	0.1431	101	-0.51	004.07	0
004	Fiber, Crude (%)	2144	5.680	0.0800	5.875	0.3893	0.1431	101	-0.50	004.07	0
004	Fiber, Crude (%)	2144	5.690	0.0200	5.875	0.3893	0.1431	101	-0.48	004.01	0
004	Fiber, Crude (%)	0164	5.700	0.0000	5.875	0.3893	0.1431	101	-0.45	004.00	0
004	Fiber, Crude (%)	0148	5.700	0.0000	5.875	0.3893	0.1431	101	-0.45	004.11	0
004	Fiber, Crude (%)	0638	5.723	0.3730	5.875	0.3893	0.1431	101	-0.39	004.06	0
004	Fiber, Crude (%)	2181	5.735	0.2300	5.875	0.3893	0.1431	101	-0.36	004.07	0
004	Fiber, Crude (%)	0708	5.740	0.0000	5.875	0.3893	0.1431	101	-0.35	004.07	0
004	Fiber, Crude (%)	2196	5.750	0.0000	5.875	0.3893	0.1431	101	-0.32	004.07	0
004	Fiber, Crude (%)	0848	5.765	0.0100	5.875	0.3893	0.1431	101	-0.28	004.07	0
004	Fiber, Crude (%)	0098	5.775	0.0300	5.875	0.3893	0.1431	101	-0.26	004.07	0
004	Fiber, Crude (%)	0885	5.784	0.1495	5.875	0.3893	0.1431	101	-0.23	004.06	0
004	Fiber, Crude (%)	0083	5.785	0.0900	5.875	0.3893	0.1431	101	-0.23	004.07	0
004	Fiber, Crude (%)	0229	5.795	0.6500	5.875	0.3893	0.1431	101	-0.21	004.07	0
004	Fiber, Crude (%)	0876	5.800	0.0000	5.875	0.3893	0.1431	101	-0.19	004.07	0
004	Fiber, Crude (%)	0581	5.800	0.2000	5.875	0.3893	0.1431	101	-0.19	004.07	0
004	Fiber, Crude (%)	0889	5.833	0.0650	5.875	0.3893	0.1431	101	-0.11	004.11	0
004	Fiber, Crude (%)	2302	5.835	0.1700	5.875	0.3893	0.1431	101	-0.10	004.07	0
004	Fiber, Crude (%)	2295	5.845	0.0100	5.875	0.3893	0.1431	101	-0.08	004.06	0
004	Fiber, Crude (%)	0353	5.845	0.0300	5.875	0.3893	0.1431	101	-0.08	004.03	0
004	Fiber, Crude (%)	2006	5.845	0.2700	5.875	0.3893	0.1431	101	-0.08	004.06	0
004	Fiber, Crude (%)	0354	5.860	0.0400	5.875	0.3893	0.1431	101	-0.04	004.07	0
004	Fiber, Crude (%)	0144	5.885	0.0900	5.875	0.3893	0.1431	101	0.03	004.07	0
004	Fiber, Crude (%)	0089	5.890	0.0400	5.875	0.3893	0.1431	101	0.04	004.07	0
004	Fiber, Crude (%)	0948	5.895	0.0300	5.875	0.3893	0.1431	101	0.05	004.06	0
004	Fiber, Crude (%)	0643	5.930	0.0200	5.875	0.3893	0.1431	101	0.14	004.07	0
004	Fiber, Crude (%)	0100	5.935	0.3700	5.875	0.3893	0.1431	101	0.15	004.07	0
004	Fiber, Crude (%)	2188	5.935	0.0500	5.875	0.3893	0.1431	101	0.15	004.07	0
004	Fiber, Crude (%)	0045	5.940	0.2400	5.875	0.3893	0.1431	101	0.17	004.07	0
004	Fiber, Crude (%)	0309	5.945	0.0083	5.875	0.3893	0.1431	101	0.18	004.00	0
004	Fiber, Crude (%)	0169	5.950	0.1000	5.875	0.3893	0.1431	101	0.19	004.00	0
004	Fiber, Crude (%)	0674	5.950	0.0000	5.875	0.3893	0.1431	101	0.19	004.06	0
004	Fiber, Crude (%)	0505	5.960	0.0400	5.875	0.3893	0.1431	101	0.22	004.07	0
004	Fiber, Crude (%)	0968	5.965	0.0080	5.875	0.3893	0.1431	101	0.23	004.06	0
004	Fiber, Crude (%)	0098	5.970	0.2400	5.875	0.3893	0.1431	101	0.24	004.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
004	Fiber, Crude (%)	0074	5.985	0.0300	5.875	0.3893	0.1431	101	0.28	004.07	0
004	Fiber, Crude (%)	0171	6.005	0.1100	5.875	0.3893	0.1431	101	0.33	004.00	0
004	Fiber, Crude (%)	0066	6.020	0.3400	5.875	0.3893	0.1431	101	0.37	004.07	0
004	Fiber, Crude (%)	0298	6.020	0.0200	5.875	0.3893	0.1431	101	0.37	004.07	0
004	Fiber, Crude (%)	2062	6.047	0.1577	5.875	0.3893	0.1431	101	0.44	004.06	0
004	Fiber, Crude (%)	0520	6.050	0.1000	5.875	0.3893	0.1431	101	0.45	004.07	0
004	Fiber, Crude (%)	0884	6.050	0.3000	5.875	0.3893	0.1431	101	0.45	004.00	0
004	Fiber, Crude (%)	0941	6.055	0.2100	5.875	0.3893	0.1431	101	0.46	004.06	0
004	Fiber, Crude (%)	0263	6.057	0.1120	5.875	0.3893	0.1431	101	0.47	004.07	0
004	Fiber, Crude (%)	0939	6.065	0.5900	5.875	0.3893	0.1431	101	0.49	004.07	0
004	Fiber, Crude (%)	0878	6.085	0.1300	5.875	0.3893	0.1431	101	0.54	004.06	0
004	Fiber, Crude (%)	0948	6.090	0.2000	5.875	0.3893	0.1431	101	0.55	004.07	0
004	Fiber, Crude (%)	0870	6.093	0.0492	5.875	0.3893	0.1431	101	0.56	004.07	0
004	Fiber, Crude (%)	0881	6.094	0.1440	5.875	0.3893	0.1431	101	0.56	004.06	0
004	Fiber, Crude (%)	0345	6.100	0.1000	5.875	0.3893	0.1431	101	0.58	004.00	0
004	Fiber, Crude (%)	0034	6.125	0.1300	5.875	0.3893	0.1431	101	0.64	004.07	0
004	Fiber, Crude (%)	0019	6.145	0.5700	5.875	0.3893	0.1431	101	0.69	004.07	0
004	Fiber, Crude (%)	0529	6.185	0.0500	5.875	0.3893	0.1431	101	0.80	004.07	0
004	Fiber, Crude (%)	0868	6.195	0.1500	5.875	0.3893	0.1431	101	0.82	004.06	0
004	Fiber, Crude (%)	0536	6.195	0.1300	5.875	0.3893	0.1431	101	0.82	004.07	0
004	Fiber, Crude (%)	2129	6.206	0.1190	5.875	0.3893	0.1431	101	0.85	004.99	0
004	Fiber, Crude (%)	2073	6.235	0.0300	5.875	0.3893	0.1431	101	0.92	004.00	0
004	Fiber, Crude (%)	0918	6.250	0.1800	5.875	0.3893	0.1431	101	0.96	004.06	0
004	Fiber, Crude (%)	0598	6.285	0.0700	5.875	0.3893	0.1431	101	1.05	004.07	0
004	Fiber, Crude (%)	0692	6.300	0.2000	5.875	0.3893	0.1431	101	1.09	004.07	0
004	Fiber, Crude (%)	2192	6.360	0.1800	5.875	0.3893	0.1431	101	1.25	004.03	0
004	Fiber, Crude (%)	0199	6.365	0.0900	5.875	0.3893	0.1431	101	1.26	004.06	0
004	Fiber, Crude (%)	2076	6.446	0.5060	5.875	0.3893	0.1431	101	1.47	004.00	0
004	Fiber, Crude (%)	0294	6.450	0.1000	5.875	0.3893	0.1431	101	1.48	004.07	0
004	Fiber, Crude (%)	2033	6.450	0.1000	5.875	0.3893	0.1431	101	1.48	004.07	0
004	Fiber, Crude (%)	0553	6.570	0.5800	5.875	0.3893	0.1431	101	1.78	004.07	0
004	Fiber, Crude (%)	0009	6.595	0.2300	5.875	0.3893	0.1431	101	1.85	004.07	0
004	Fiber, Crude (%)	0208	6.605	0.3900	5.875	0.3893	0.1431	101	1.87	004.00	0
004	Fiber, Crude (%)	0032	6.750	0.1000	5.875	0.3893	0.1431	101	2.25	004.07	0
004	Fiber, Crude (%)	0226	7.120	0.0800	5.875	0.3893	0.1431	101	3.20	004.00	0
004	Fiber, Crude (%)	0626	7.190	0.1400	5.875	0.3893	0.1431	101	3.38	004.07	0
004	Fiber, Crude (%)	0278	8.300	0.0000	5.875	0.3893	0.1431	101	6.23	004.07	0
004	Fiber, Crude (%)	2291	8.445	0.4100	5.875	0.3893	0.1431	101	6.60	004.06	0
004	Fiber, Crude (%)	0646	5.315	1.250	5.875	0.3893	0.1431	101	-1.44	004.07	1
004	Fiber, Crude (%)	0693	6.777	1.095	5.875	0.3893	0.1431	101	2.31	004.07	1
005	Ash (%)	0148	7.100	0.0000	8.149	0.2511	0.0740	111	-4.17	005.11	0
005	Ash (%)	2081	7.375	0.2900	8.149	0.2511	0.0740	111	-3.08	005.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
005	Ash (%)	0278	7.700	0.0600	8.149	0.2511	0.0740	111	-1.79	005.00	0
005	Ash (%)	0144	7.710	0.0800	8.149	0.2511	0.0740	111	-1.75	005.00	0
005	Ash (%)	2184	7.745	0.2300	8.149	0.2511	0.0740	111	-1.61	005.00	0
005	Ash (%)	0019	7.750	0.1000	8.149	0.2511	0.0740	111	-1.59	005.00	0
005	Ash (%)	0169	7.750	0.1000	8.149	0.2511	0.0740	111	-1.59	005.00	0
005	Ash (%)	2033	7.750	0.1000	8.149	0.2511	0.0740	111	-1.59	005.00	0
005	Ash (%)	0062	7.764	0.1000	8.149	0.2511	0.0740	111	-1.53	005.00	0
005	Ash (%)	0357	7.800	0.0000	8.149	0.2511	0.0740	111	-1.39	005.00	0
005	Ash (%)	0726	7.840	0.0600	8.149	0.2511	0.0740	111	-1.23	005.05	0
005	Ash (%)	0953	7.850	0.1600	8.149	0.2511	0.0740	111	-1.19	005.00	0
005	Ash (%)	0574	7.860	0.0800	8.149	0.2511	0.0740	111	-1.15	005.00	0
005	Ash (%)	0870	7.867	0.0391	8.149	0.2511	0.0740	111	-1.12	005.00	0
005	Ash (%)	0541	7.870	0.0400	8.149	0.2511	0.0740	111	-1.11	005.00	0
005	Ash (%)	0208	7.880	0.1200	8.149	0.2511	0.0740	111	-1.07	005.00	0
005	Ash (%)	2073	7.890	0.0600	8.149	0.2511	0.0740	111	-1.03	005.00	0
005	Ash (%)	0297	7.890	0.1800	8.149	0.2511	0.0740	111	-1.03	005.11	0
005	Ash (%)	0175	7.920	0.0400	8.149	0.2511	0.0740	111	-0.91	005.00	0
005	Ash (%)	2304	7.920	0.0000	8.149	0.2511	0.0740	111	-0.91	005.03	0
005	Ash (%)	0589	7.940	0.3800	8.149	0.2511	0.0740	111	-0.83	005.00	0
005	Ash (%)	0682	7.940	0.0000	8.149	0.2511	0.0740	111	-0.83	005.00	0
005	Ash (%)	2302	7.950	0.1000	8.149	0.2511	0.0740	111	-0.79	005.00	0
005	Ash (%)	0407	7.955	0.0443	8.149	0.2511	0.0740	111	-0.77	005.00	0
005	Ash (%)	0598	7.965	0.0500	8.149	0.2511	0.0740	111	-0.73	005.00	0
005	Ash (%)	0918	7.970	0.0600	8.149	0.2511	0.0740	111	-0.71	005.00	0
005	Ash (%)	0066	7.980	0.0600	8.149	0.2511	0.0740	111	-0.67	005.05	0
005	Ash (%)	0510	7.985	0.1300	8.149	0.2511	0.0740	111	-0.65	005.00	0
005	Ash (%)	0142	7.990	0.0400	8.149	0.2511	0.0740	111	-0.63	005.00	0
005	Ash (%)	2188	7.990	0.0000	8.149	0.2511	0.0740	111	-0.63	005.00	0
005	Ash (%)	0529	8.000	0.0000	8.149	0.2511	0.0740	111	-0.59	005.00	0
005	Ash (%)	0646	8.015	0.0700	8.149	0.2511	0.0740	111	-0.53	005.00	0
005	Ash (%)	0164	8.020	0.0400	8.149	0.2511	0.0740	111	-0.51	005.00	0
005	Ash (%)	0553	8.020	0.0800	8.149	0.2511	0.0740	111	-0.51	005.00	0
005	Ash (%)	0003	8.030	0.0000	8.149	0.2511	0.0740	111	-0.47	005.05	0
005	Ash (%)	2053	8.035	0.0300	8.149	0.2511	0.0740	111	-0.45	005.00	0
005	Ash (%)	0728	8.043	0.0150	8.149	0.2511	0.0740	111	-0.42	005.99	0
005	Ash (%)	2150	8.060	0.1000	8.149	0.2511	0.0740	111	-0.35	005.00	0
005	Ash (%)	0968	8.060	0.0760	8.149	0.2511	0.0740	111	-0.35	005.00	0
005	Ash (%)	0358	8.065	0.0900	8.149	0.2511	0.0740	111	-0.33	005.00	0
005	Ash (%)	0693	8.082	0.0620	8.149	0.2511	0.0740	111	-0.26	005.00	0
005	Ash (%)	0263	8.083	0.0180	8.149	0.2511	0.0740	111	-0.26	005.05	0
005	Ash (%)	0171	8.085	0.0100	8.149	0.2511	0.0740	111	-0.25	005.00	0
005	Ash (%)	0265	8.100	0.0000	8.149	0.2511	0.0740	111	-0.19	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
005	Ash (%)	0186	8.103	0.0720	8.149	0.2511	0.0740	111	-0.18	005.00	0
005	Ash (%)	0229	8.105	0.3900	8.149	0.2511	0.0740	111	-0.17	005.00	0
005	Ash (%)	0964	8.110	0.0280	8.149	0.2511	0.0740	111	-0.15	005.00	0
005	Ash (%)	2012	8.110	0.0400	8.149	0.2511	0.0740	111	-0.15	005.05	0
005	Ash (%)	0881	8.119	0.0480	8.149	0.2511	0.0740	111	-0.12	005.00	0
005	Ash (%)	0098	8.130	0.1400	8.149	0.2511	0.0740	111	-0.07	005.00	0
005	Ash (%)	0656	8.130	0.0000	8.149	0.2511	0.0740	111	-0.07	005.00	0
005	Ash (%)	0001	8.137	0.0665	8.149	0.2511	0.0740	111	-0.05	005.00	0
005	Ash (%)	0100	8.140	0.1800	8.149	0.2511	0.0740	111	-0.03	005.00	0
005	Ash (%)	0868	8.140	0.0000	8.149	0.2511	0.0740	111	-0.03	005.05	0
005	Ash (%)	2291	8.140	0.0000	8.149	0.2511	0.0740	111	-0.03	005.05	0
005	Ash (%)	0035	8.145	0.0100	8.149	0.2511	0.0740	111	-0.01	005.00	0
005	Ash (%)	0083	8.145	0.0500	8.149	0.2511	0.0740	111	-0.01	005.00	0
005	Ash (%)	0683	8.145	0.0900	8.149	0.2511	0.0740	111	-0.01	005.05	0
005	Ash (%)	0876	8.150	0.1000	8.149	0.2511	0.0740	111	0.01	005.00	0
005	Ash (%)	0692	8.150	0.1000	8.149	0.2511	0.0740	111	0.01	005.05	0
005	Ash (%)	2089	8.150	0.0400	8.149	0.2511	0.0740	111	0.01	005.00	0
005	Ash (%)	2196	8.150	0.0000	8.149	0.2511	0.0740	111	0.01	005.00	0
005	Ash (%)	0643	8.160	0.0400	8.149	0.2511	0.0740	111	0.05	005.00	0
005	Ash (%)	0353	8.170	0.0200	8.149	0.2511	0.0740	111	0.09	005.00	0
005	Ash (%)	2006	8.170	0.0800	8.149	0.2511	0.0740	111	0.09	005.00	0
005	Ash (%)	0298	8.180	0.0200	8.149	0.2511	0.0740	111	0.13	005.00	0
005	Ash (%)	0354	8.180	0.0000	8.149	0.2511	0.0740	111	0.13	005.00	0
005	Ash (%)	0848	8.180	0.0200	8.149	0.2511	0.0740	111	0.13	005.00	0
005	Ash (%)	2076	8.194	0.0182	8.149	0.2511	0.0740	111	0.18	005.00	0
005	Ash (%)	0045	8.195	0.1700	8.149	0.2511	0.0740	111	0.19	005.00	0
005	Ash (%)	0199	8.200	0.2000	8.149	0.2511	0.0740	111	0.21	005.00	0
005	Ash (%)	0968	8.215	0.0080	8.149	0.2511	0.0740	111	0.26	005.02	0
005	Ash (%)	2192	8.215	0.0500	8.149	0.2511	0.0740	111	0.26	005.05	0
005	Ash (%)	0674	8.230	0.0000	8.149	0.2511	0.0740	111	0.32	005.00	0
005	Ash (%)	0226	8.240	0.0400	8.149	0.2511	0.0740	111	0.36	005.00	0
005	Ash (%)	0914	8.240	0.1600	8.149	0.2511	0.0740	111	0.36	005.00	0
005	Ash (%)	0853	8.250	0.1000	8.149	0.2511	0.0740	111	0.40	005.00	0
005	Ash (%)	0948	8.260	0.0000	8.149	0.2511	0.0740	111	0.44	005.00	0
005	Ash (%)	2181	8.260	0.0200	8.149	0.2511	0.0740	111	0.44	005.00	0
005	Ash (%)	0536	8.260	0.0200	8.149	0.2511	0.0740	111	0.44	005.99	0
005	Ash (%)	0505	8.285	0.0700	8.149	0.2511	0.0740	111	0.54	005.00	0
005	Ash (%)	0034	8.295	0.1630	8.149	0.2511	0.0740	111	0.58	005.05	0
005	Ash (%)	0956	8.300	0.2000	8.149	0.2511	0.0740	111	0.60	005.00	0
005	Ash (%)	2109	8.305	0.1100	8.149	0.2511	0.0740	111	0.62	005.05	0
005	Ash (%)	0878	8.320	0.0800	8.149	0.2511	0.0740	111	0.68	005.05	0
005	Ash (%)	0202	8.355	0.0300	8.149	0.2511	0.0740	111	0.82	005.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
005	Ash (%)	2062	8.361	0.0066	8.149	0.2511	0.0740	111	0.85	005.05	0
005	Ash (%)	0520	8.370	0.0600	8.149	0.2511	0.0740	111	0.88	005.00	0
005	Ash (%)	0939	8.370	0.1200	8.149	0.2511	0.0740	111	0.88	005.05	0
005	Ash (%)	0889	8.385	0.1100	8.149	0.2511	0.0740	111	0.94	005.05	0
005	Ash (%)	0675	8.410	0.1400	8.149	0.2511	0.0740	111	1.04	005.00	0
005	Ash (%)	0676	8.410	0.3600	8.149	0.2511	0.0740	111	1.04	005.99	0
005	Ash (%)	0123	8.425	0.1100	8.149	0.2511	0.0740	111	1.10	005.00	0
005	Ash (%)	0148	8.430	0.0600	8.149	0.2511	0.0740	111	1.12	005.00	0
005	Ash (%)	2146	8.435	0.0100	8.149	0.2511	0.0740	111	1.14	005.05	0
005	Ash (%)	0309	8.442	0.1090	8.149	0.2511	0.0740	111	1.17	005.00	0
005	Ash (%)	0884	8.450	0.1000	8.149	0.2511	0.0740	111	1.20	005.99	0
005	Ash (%)	0139	8.465	0.0100	8.149	0.2511	0.0740	111	1.26	005.00	0
005	Ash (%)	0590	8.470	0.0400	8.149	0.2511	0.0740	111	1.28	005.05	0
005	Ash (%)	2295	8.475	0.0100	8.149	0.2511	0.0740	111	1.30	005.05	0
005	Ash (%)	0345	8.485	0.0500	8.149	0.2511	0.0740	111	1.34	005.00	0
005	Ash (%)	0885	8.488	0.1914	8.149	0.2511	0.0740	111	1.35	005.05	0
005	Ash (%)	0294	8.500	0.2000	8.149	0.2511	0.0740	111	1.40	005.05	0
005	Ash (%)	0652	8.500	0.0000	8.149	0.2511	0.0740	111	1.40	005.99	0
005	Ash (%)	2259	8.511	0.0210	8.149	0.2511	0.0740	111	1.44	005.00	0
005	Ash (%)	2129	8.568	0.0960	8.149	0.2511	0.0740	111	1.67	005.05	0
005	Ash (%)	0921	8.575	0.0300	8.149	0.2511	0.0740	111	1.70	005.05	0
005	Ash (%)	0910	8.600	0.0000	8.149	0.2511	0.0740	111	1.80	005.99	0
005	Ash (%)	0190	8.715	0.0300	8.149	0.2511	0.0740	111	2.26	005.05	0
005	Ash (%)	0581	8.830	0.0800	8.149	0.2511	0.0740	111	2.71	005.00	0
005	Ash (%)	0425	9.620	0.0000	8.149	0.2511	0.0740	111	5.86	005.00	0
005	Ash (%)	0638	7.578	0.9820	8.149	0.2511	0.0740	111	-2.27	005.05	1
005	Ash (%)	0297	8.190	0.7800	8.149	0.2511	0.0740	111	0.17	005.00	1
005	Ash (%)	0618	9.212	1.478	8.149	0.2511	0.0740	111	4.24	005.00	1
005	Ash (%)	0889	10.57	0.0100	8.149	0.2511	0.0740	111	9.62	005.11	2
006	Total Sugars (%)	0921	4.915	0.1300	5.650	0.7683	0.1160	4		006.00	0
006	Total Sugars (%)	0407	5.294	0.1340	5.650	0.7683	0.1160	4		006.00	0
006	Total Sugars (%)	0910	5.690	0.0000	5.650	0.7683	0.1160	4		006.00	0
006	Total Sugars (%)	0956	6.700	0.2000	5.650	0.7683	0.1160	4		006.99	0
008	Fiber, Acid Detergent (%)	0038	6.232	0.3610	7.717	0.6573	0.1846	48	-2.26	008.02	0
008	Fiber, Acid Detergent (%)	0956	6.550	0.7000	7.717	0.6573	0.1846	48	-1.77	008.08	0
008	Fiber, Acid Detergent (%)	2302	6.600	0.0800	7.717	0.6573	0.1846	48	-1.70	008.08	0
008	Fiber, Acid Detergent (%)	0968	6.628	0.0850	7.717	0.6573	0.1846	48	-1.66	008.08	0
008	Fiber, Acid Detergent (%)	0590	6.870	0.5000	7.717	0.6573	0.1846	48	-1.29	008.08	0
008	Fiber, Acid Detergent (%)	0939	6.995	0.0700	7.717	0.6573	0.1846	48	-1.10	008.08	0
008	Fiber, Acid Detergent (%)	0675	7.090	0.0600	7.717	0.6573	0.1846	48	-0.95	008.08	0
008	Fiber, Acid Detergent (%)	0910	7.110	0.0000	7.717	0.6573	0.1846	48	-0.92	008.08	0
008	Fiber, Acid Detergent (%)	0948	7.115	0.0900	7.717	0.6573	0.1846	48	-0.92	008.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
008	Fiber, Acid Detergent (%)	2012	7.130	0.0200	7.717	0.6573	0.1846	48	-0.89	008.08	0
008	Fiber, Acid Detergent (%)	0098	7.235	0.2100	7.717	0.6573	0.1846	48	-0.73	008.02	0
008	Fiber, Acid Detergent (%)	0878	7.250	0.1000	7.717	0.6573	0.1846	48	-0.71	008.08	0
008	Fiber, Acid Detergent (%)	0848	7.280	0.1400	7.717	0.6573	0.1846	48	-0.66	008.08	0
008	Fiber, Acid Detergent (%)	0918	7.405	0.3575	7.717	0.6573	0.1846	48	-0.47	008.08	0
008	Fiber, Acid Detergent (%)	0358	7.420	0.3400	7.717	0.6573	0.1846	48	-0.45	008.08	0
008	Fiber, Acid Detergent (%)	0726	7.515	0.1300	7.717	0.6573	0.1846	48	-0.31	008.08	0
008	Fiber, Acid Detergent (%)	0353	7.545	0.1100	7.717	0.6573	0.1846	48	-0.26	008.02	0
008	Fiber, Acid Detergent (%)	0294	7.550	0.1000	7.717	0.6573	0.1846	48	-0.25	008.08	0
008	Fiber, Acid Detergent (%)	0190	7.570	0.1200	7.717	0.6573	0.1846	48	-0.22	008.08	0
008	Fiber, Acid Detergent (%)	0870	7.571	0.0507	7.717	0.6573	0.1846	48	-0.22	008.08	0
008	Fiber, Acid Detergent (%)	0510	7.600	0.0000	7.717	0.6573	0.1846	48	-0.18	008.08	0
008	Fiber, Acid Detergent (%)	0083	7.605	0.2300	7.717	0.6573	0.1846	48	-0.17	008.08	0
008	Fiber, Acid Detergent (%)	0297	7.610	0.0200	7.717	0.6573	0.1846	48	-0.16	008.08	0
008	Fiber, Acid Detergent (%)	0948	7.619	0.1130	7.717	0.6573	0.1846	48	-0.15	008.08	0
008	Fiber, Acid Detergent (%)	0164	7.650	0.3000	7.717	0.6573	0.1846	48	-0.10	008.08	0
008	Fiber, Acid Detergent (%)	0357	7.650	0.1000	7.717	0.6573	0.1846	48	-0.10	008.08	0
008	Fiber, Acid Detergent (%)	0345	7.700	0.2000	7.717	0.6573	0.1846	48	-0.03	008.02	0
008	Fiber, Acid Detergent (%)	0001	7.860	0.2600	7.717	0.6573	0.1846	48	0.22	008.08	0
008	Fiber, Acid Detergent (%)	0309	7.875	0.0087	7.717	0.6573	0.1846	48	0.24	008.02	0
008	Fiber, Acid Detergent (%)	0148	7.915	0.0500	7.717	0.6573	0.1846	48	0.30	008.08	0
008	Fiber, Acid Detergent (%)	0265	7.970	0.1600	7.717	0.6573	0.1846	48	0.39	008.05	0
008	Fiber, Acid Detergent (%)	0148	7.985	0.3300	7.717	0.6573	0.1846	48	0.41	008.02	0
008	Fiber, Acid Detergent (%)	0278	8.000	0.2000	7.717	0.6573	0.1846	48	0.43	008.08	0
008	Fiber, Acid Detergent (%)	0581	8.000	0.2000	7.717	0.6573	0.1846	48	0.43	008.08	0
008	Fiber, Acid Detergent (%)	0066	8.005	0.3300	7.717	0.6573	0.1846	48	0.44	008.08	0
008	Fiber, Acid Detergent (%)	2196	8.010	0.0000	7.717	0.6573	0.1846	48	0.45	008.08	0
008	Fiber, Acid Detergent (%)	0035	8.040	0.1600	7.717	0.6573	0.1846	48	0.49	008.08	0
008	Fiber, Acid Detergent (%)	0263	8.250	0.1630	7.717	0.6573	0.1846	48	0.81	008.08	0
008	Fiber, Acid Detergent (%)	2192	8.255	0.1500	7.717	0.6573	0.1846	48	0.82	008.02	0
008	Fiber, Acid Detergent (%)	0407	8.301	0.1042	7.717	0.6573	0.1846	48	0.89	008.08	0
008	Fiber, Acid Detergent (%)	0354	8.335	0.0100	7.717	0.6573	0.1846	48	0.94	008.08	0
008	Fiber, Acid Detergent (%)	0646	8.455	0.0100	7.717	0.6573	0.1846	48	1.12	008.08	0
008	Fiber, Acid Detergent (%)	0045	8.630	0.3200	7.717	0.6573	0.1846	48	1.39	008.08	0
008	Fiber, Acid Detergent (%)	0226	8.695	0.0500	7.717	0.6573	0.1846	48	1.49	008.02	0
008	Fiber, Acid Detergent (%)	0037	8.740	0.4000	7.717	0.6573	0.1846	48	1.56	008.08	0
008	Fiber, Acid Detergent (%)	0693	8.798	0.6600	7.717	0.6573	0.1846	48	1.65	008.08	0
008	Fiber, Acid Detergent (%)	2129	8.939	0.3400	7.717	0.6573	0.1846	48	1.86	008.08	0
008	Fiber, Acid Detergent (%)	0536	9.665	0.3700	7.717	0.6573	0.1846	48	2.96	008.08	0
008	Fiber, Acid Detergent (%)	0964	19.60	0.1800	7.717	0.6573	0.1846	48	18.08	008.02	2
009	Fiber, Neutral Detergent (%)	0910	12.70	0.0000	14.60	0.8756	0.3055	48	-2.17	009.07	0
009	Fiber, Neutral Detergent (%)	0939	12.98	0.3000	14.60	0.8756	0.3055	48	-1.85	009.09	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
009	Fiber, Neutral Detergent (%)	0878	13.37	0.1200	14.60	0.8756	0.3055	48	-1.40	009.09	0
009	Fiber, Neutral Detergent (%)	0358	13.38	0.5200	14.60	0.8756	0.3055	48	-1.39	009.09	0
009	Fiber, Neutral Detergent (%)	0598	13.63	0.1500	14.60	0.8756	0.3055	48	-1.11	009.09	0
009	Fiber, Neutral Detergent (%)	0035	13.66	0.1600	14.60	0.8756	0.3055	48	-1.07	009.09	0
009	Fiber, Neutral Detergent (%)	0353	13.67	0.1300	14.60	0.8756	0.3055	48	-1.07	009.07	0
009	Fiber, Neutral Detergent (%)	0675	13.72	0.0700	14.60	0.8756	0.3055	48	-1.01	009.09	0
009	Fiber, Neutral Detergent (%)	0083	13.75	0.1400	14.60	0.8756	0.3055	48	-0.97	009.09	0
009	Fiber, Neutral Detergent (%)	0263	13.76	0.1190	14.60	0.8756	0.3055	48	-0.96	009.09	0
009	Fiber, Neutral Detergent (%)	0164	13.80	0.4000	14.60	0.8756	0.3055	48	-0.91	009.09	0
009	Fiber, Neutral Detergent (%)	2302	13.93	0.3500	14.60	0.8756	0.3055	48	-0.77	009.09	0
009	Fiber, Neutral Detergent (%)	0956	13.95	1.100	14.60	0.8756	0.3055	48	-0.74	009.09	0
009	Fiber, Neutral Detergent (%)	0294	14.15	0.1000	14.60	0.8756	0.3055	48	-0.51	009.09	0
009	Fiber, Neutral Detergent (%)	0510	14.15	0.3000	14.60	0.8756	0.3055	48	-0.51	009.09	0
009	Fiber, Neutral Detergent (%)	0870	14.19	0.1179	14.60	0.8756	0.3055	48	-0.46	009.09	0
009	Fiber, Neutral Detergent (%)	0148	14.25	0.1000	14.60	0.8756	0.3055	48	-0.40	009.09	0
009	Fiber, Neutral Detergent (%)	0357	14.40	0.4000	14.60	0.8756	0.3055	48	-0.23	009.09	0
009	Fiber, Neutral Detergent (%)	0581	14.40	0.2000	14.60	0.8756	0.3055	48	-0.23	009.09	0
009	Fiber, Neutral Detergent (%)	0646	14.47	0.1000	14.60	0.8756	0.3055	48	-0.15	009.09	0
009	Fiber, Neutral Detergent (%)	0948	14.48	0.3000	14.60	0.8756	0.3055	48	-0.13	009.07	0
009	Fiber, Neutral Detergent (%)	0968	14.49	0.0870	14.60	0.8756	0.3055	48	-0.13	009.09	0
009	Fiber, Neutral Detergent (%)	0098	14.50	0.4000	14.60	0.8756	0.3055	48	-0.11	009.07	0
009	Fiber, Neutral Detergent (%)	0066	14.53	0.3000	14.60	0.8756	0.3055	48	-0.08	009.09	0
009	Fiber, Neutral Detergent (%)	0037	14.54	0.5200	14.60	0.8756	0.3055	48	-0.07	009.09	0
009	Fiber, Neutral Detergent (%)	0265	14.55	0.3000	14.60	0.8756	0.3055	48	-0.06	009.09	0
009	Fiber, Neutral Detergent (%)	0407	14.74	0.7525	14.60	0.8756	0.3055	48	0.17	009.09	0
009	Fiber, Neutral Detergent (%)	0948	14.79	0.1470	14.60	0.8756	0.3055	48	0.22	009.09	0
009	Fiber, Neutral Detergent (%)	2196	14.80	0.0000	14.60	0.8756	0.3055	48	0.23	009.09	0
009	Fiber, Neutral Detergent (%)	0881	14.83	0.7520	14.60	0.8756	0.3055	48	0.26	009.09	0
009	Fiber, Neutral Detergent (%)	2129	14.87	0.7210	14.60	0.8756	0.3055	48	0.31	009.09	0
009	Fiber, Neutral Detergent (%)	2012	14.90	0.0800	14.60	0.8756	0.3055	48	0.34	009.09	0
009	Fiber, Neutral Detergent (%)	0693	15.00	0.3070	14.60	0.8756	0.3055	48	0.46	009.09	0
009	Fiber, Neutral Detergent (%)	0345	15.00	0.4000	14.60	0.8756	0.3055	48	0.46	009.07	0
009	Fiber, Neutral Detergent (%)	0590	15.10	0.0000	14.60	0.8756	0.3055	48	0.57	009.09	0
009	Fiber, Neutral Detergent (%)	0726	15.14	0.4200	14.60	0.8756	0.3055	48	0.62	009.09	0
009	Fiber, Neutral Detergent (%)	0918	15.15	0.0234	14.60	0.8756	0.3055	48	0.63	009.09	0
009	Fiber, Neutral Detergent (%)	0297	15.29	0.9800	14.60	0.8756	0.3055	48	0.79	009.09	0
009	Fiber, Neutral Detergent (%)	0045	15.30	0.8000	14.60	0.8756	0.3055	48	0.80	009.09	0
009	Fiber, Neutral Detergent (%)	0278	15.30	0.2000	14.60	0.8756	0.3055	48	0.80	009.09	0
009	Fiber, Neutral Detergent (%)	0536	15.53	0.2300	14.60	0.8756	0.3055	48	1.06	009.09	0
009	Fiber, Neutral Detergent (%)	0226	15.55	0.5400	14.60	0.8756	0.3055	48	1.09	009.07	0
009	Fiber, Neutral Detergent (%)	0889	15.62	0.0800	14.60	0.8756	0.3055	48	1.17	009.99	0
009	Fiber, Neutral Detergent (%)	0848	15.90	0.1500	14.60	0.8756	0.3055	48	1.48	009.09	0



Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
009	Fiber, Neutral Detergent (%)	2192	16.37	0.6600	14.60	0.8756	0.3055	48	2.02	009.07	0
009	Fiber, Neutral Detergent (%)	0354	16.67	0.0800	14.60	0.8756	0.3055	48	2.37	009.09	0
009	Fiber, Neutral Detergent (%)	0190	16.87	0.5000	14.60	0.8756	0.3055	48	2.59	009.09	0
009	Fiber, Neutral Detergent (%)	0309	17.01	0.0562	14.60	0.8756	0.3055	48	2.75	009.07	0
010	Moisture (%)	0843	8.510	0.6400	9.861	0.6114	0.1007	19	-2.21	010.03	0
010	Moisture (%)	0964	8.543	0.0380	9.861	0.6114	0.1007	19	-2.16	010.99	0
010	Moisture (%)	0889	9.075	0.0900	9.861	0.6114	0.1007	19	-1.29	010.11	0
010	Moisture (%)	0726	9.260	0.0000	9.861	0.6114	0.1007	19	-0.98	010.99	0
010	Moisture (%)	0948	9.490	0.0000	9.861	0.6114	0.1007	19	-0.61	010.99	0
010	Moisture (%)	0939	9.635	0.0100	9.861	0.6114	0.1007	19	-0.37	010.99	0
010	Moisture (%)	0652	9.700	0.0000	9.861	0.6114	0.1007	19	-0.26	010.99	0
010	Moisture (%)	2089	9.760	0.0200	9.861	0.6114	0.1007	19	-0.16	010.99	0
010	Moisture (%)	0148	9.785	0.0300	9.861	0.6114	0.1007	19	-0.12	010.11	0
010	Moisture (%)	2150	9.990	0.1400	9.861	0.6114	0.1007	19	0.21	010.99	0
010	Moisture (%)	0884	10.00	0.0000	9.861	0.6114	0.1007	19	0.23	010.99	0
010	Moisture (%)	0164	10.02	0.0700	9.861	0.6114	0.1007	19	0.25	010.03	0
010	Moisture (%)	2129	10.14	0.1760	9.861	0.6114	0.1007	19	0.46	010.99	0
010	Moisture (%)	0921	10.21	0.0500	9.861	0.6114	0.1007	19	0.56	010.99	0
010	Moisture (%)	2076	10.29	0.1800	9.861	0.6114	0.1007	19	0.70	010.99	0
010	Moisture (%)	2178	10.33	0.3600	9.861	0.6114	0.1007	19	0.77	010.99	0
010	Moisture (%)	0190	10.47	0.0200	9.861	0.6114	0.1007	19	1.00	010.99	0
010	Moisture (%)	0914	10.55	0.0300	9.861	0.6114	0.1007	19	1.12	010.99	0
010	Moisture (%)	2302	11.12	0.0600	9.861	0.6114	0.1007	19	2.06	010.11	0
010	Moisture (%)	0692	9.720	0.7600	9.861	0.6114	0.1007	19	-0.23	010.99	1
011	Loss on Drying (%)	0520	9.300	0.0000	10.92	0.4674	0.1054	57	-3.47	011.01	0
011	Loss on Drying (%)	0536	9.810	0.0000	10.92	0.4674	0.1054	57	-2.38	011.01	0
011	Loss on Drying (%)	0062	9.817	0.1130	10.92	0.4674	0.1054	57	-2.36	011.01	0
011	Loss on Drying (%)	0175	10.00	0.2000	10.92	0.4674	0.1054	57	-1.97	011.01	0
011	Loss on Drying (%)	0643	10.25	0.5000	10.92	0.4674	0.1054	57	-1.44	011.01	0
011	Loss on Drying (%)	0425	10.28	0.4500	10.92	0.4674	0.1054	57	-1.38	011.01	0
011	Loss on Drying (%)	0294	10.30	0.0000	10.92	0.4674	0.1054	57	-1.33	011.01	0
011	Loss on Drying (%)	2303	10.32	0.2150	10.92	0.4674	0.1054	57	-1.28	011.99	0
011	Loss on Drying (%)	0354	10.48	0.0700	10.92	0.4674	0.1054	57	-0.96	011.01	0
011	Loss on Drying (%)	2033	10.49	0.0000	10.92	0.4674	0.1054	57	-0.92	011.01	0
011	Loss on Drying (%)	0598	10.50	0.0600	10.92	0.4674	0.1054	57	-0.90	011.01	0
011	Loss on Drying (%)	0674	10.63	0.0000	10.92	0.4674	0.1054	57	-0.62	011.01	0
011	Loss on Drying (%)	0942	10.64	0.3400	10.92	0.4674	0.1054	57	-0.60	011.02	0
011	Loss on Drying (%)	0859	10.64	0.0080	10.92	0.4674	0.1054	57	-0.59	011.01	0
011	Loss on Drying (%)	2168	10.66	0.0500	10.92	0.4674	0.1054	57	-0.57	011.01	0
011	Loss on Drying (%)	0263	10.66	0.0980	10.92	0.4674	0.1054	57	-0.56	011.01	0
011	Loss on Drying (%)	0953	10.72	0.0500	10.92	0.4674	0.1054	57	-0.44	011.01	0
011	Loss on Drying (%)	0881	10.79	0.4180	10.92	0.4674	0.1054	57	-0.29	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
011	Loss on Drying (%)	2291	10.79	0.1200	10.92	0.4674	0.1054	57	-0.28	011.01	0
011	Loss on Drying (%)	0968	10.80	0.0280	10.92	0.4674	0.1054	57	-0.27	011.01	0
011	Loss on Drying (%)	0208	10.80	0.0000	10.92	0.4674	0.1054	57	-0.26	011.01	0
011	Loss on Drying (%)	2174	10.80	0.1200	10.92	0.4674	0.1054	57	-0.26	011.01	0
011	Loss on Drying (%)	0728	10.80	0.1390	10.92	0.4674	0.1054	57	-0.25	011.01	0
011	Loss on Drying (%)	0646	10.82	0.1800	10.92	0.4674	0.1054	57	-0.22	011.01	0
011	Loss on Drying (%)	0098	10.83	0.1000	10.92	0.4674	0.1054	57	-0.20	011.01	0
011	Loss on Drying (%)	0553	10.84	0.2600	10.92	0.4674	0.1054	57	-0.17	011.01	0
011	Loss on Drying (%)	0226	10.85	0.3800	10.92	0.4674	0.1054	57	-0.15	011.01	0
011	Loss on Drying (%)	2196	10.85	0.0000	10.92	0.4674	0.1054	57	-0.15	011.01	0
011	Loss on Drying (%)	0510	10.85	0.1000	10.92	0.4674	0.1054	57	-0.15	011.01	0
011	Loss on Drying (%)	0100	10.89	0.3200	10.92	0.4674	0.1054	57	-0.07	011.01	0
011	Loss on Drying (%)	0265	10.90	0.0000	10.92	0.4674	0.1054	57	-0.05	011.01	0
011	Loss on Drying (%)	0298	10.90	0.0200	10.92	0.4674	0.1054	57	-0.05	011.01	0
011	Loss on Drying (%)	2006	10.91	0.2000	10.92	0.4674	0.1054	57	-0.02	011.01	0
011	Loss on Drying (%)	0574	10.96	0.0300	10.92	0.4674	0.1054	57	0.07	011.01	0
011	Loss on Drying (%)	0529	11.00	0.0200	10.92	0.4674	0.1054	57	0.17	011.02	0
011	Loss on Drying (%)	0541	11.07	0.0400	10.92	0.4674	0.1054	57	0.32	011.01	0
011	Loss on Drying (%)	2073	11.12	0.0100	10.92	0.4674	0.1054	57	0.41	011.01	0
011	Loss on Drying (%)	0682	11.14	0.0000	10.92	0.4674	0.1054	57	0.47	011.01	0
011	Loss on Drying (%)	0229	11.17	0.0900	10.92	0.4674	0.1054	57	0.52	011.01	0
011	Loss on Drying (%)	0066	11.18	0.1500	10.92	0.4674	0.1054	57	0.54	011.01	0
011	Loss on Drying (%)	0171	11.19	0.0300	10.92	0.4674	0.1054	57	0.56	011.01	0
011	Loss on Drying (%)	2181	11.20	0.0900	10.92	0.4674	0.1054	57	0.58	011.01	0
011	Loss on Drying (%)	0164	11.24	0.0600	10.92	0.4674	0.1054	57	0.68	011.01	0
011	Loss on Drying (%)	0870	11.25	0.1087	10.92	0.4674	0.1054	57	0.70	011.01	0
011	Loss on Drying (%)	2302	11.25	0.0600	10.92	0.4674	0.1054	57	0.70	011.01	0
011	Loss on Drying (%)	0123	11.27	0.0300	10.92	0.4674	0.1054	57	0.73	011.01	0
011	Loss on Drying (%)	0144	11.34	0.0900	10.92	0.4674	0.1054	57	0.88	011.01	0
011	Loss on Drying (%)	0675	11.38	0.0100	10.92	0.4674	0.1054	57	0.97	011.01	0
011	Loss on Drying (%)	0843	11.43	0.1200	10.92	0.4674	0.1054	57	1.09	011.01	0
011	Loss on Drying (%)	0358	11.46	0.0500	10.92	0.4674	0.1054	57	1.14	011.01	0
011	Loss on Drying (%)	0589	11.51	0.0300	10.92	0.4674	0.1054	57	1.25	011.01	0
011	Loss on Drying (%)	0407	11.55	0.0808	10.92	0.4674	0.1054	57	1.35	011.01	0
011	Loss on Drying (%)	0618	11.57	0.0100	10.92	0.4674	0.1054	57	1.38	011.01	0
011	Loss on Drying (%)	2184	11.59	0.2000	10.92	0.4674	0.1054	57	1.43	011.01	0
011	Loss on Drying (%)	0309	11.66	0.1200	10.92	0.4674	0.1054	57	1.58	011.01	0
011	Loss on Drying (%)	0202	11.75	0.0400	10.92	0.4674	0.1054	57	1.77	011.01	0
011	Loss on Drying (%)	2188	11.86	0.0000	10.92	0.4674	0.1054	57	2.01	011.01	0
011	Loss on Drying (%)	0857	10.15	0.7000	10.92	0.4674	0.1054	57	-1.65	011.99	1
012	Starch (%)	0510	28.80	0.2000	32.63	1.645	0.4713	32	-2.33	012.04	0
012	Starch (%)	0676	29.40	0.4700	32.63	1.645	0.4713	32	-1.97	012.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
012	Starch (%)	0265	30.00	0.8000	32.63	1.645	0.4713	32	-1.60	012.01	0
012	Starch (%)	0098	30.30	0.4000	32.63	1.645	0.4713	32	-1.42	012.04	0
012	Starch (%)	0297	30.95	0.3100	32.63	1.645	0.4713	32	-1.03	012.03	0
012	Starch (%)	0164	31.40	0.6000	32.63	1.645	0.4713	32	-0.75	012.01	0
012	Starch (%)	0918	31.45	0.1590	32.63	1.645	0.4713	32	-0.72	012.01	0
012	Starch (%)	0956	31.55	0.3000	32.63	1.645	0.4713	32	-0.66	012.01	0
012	Starch (%)	0148	31.75	0.7000	32.63	1.645	0.4713	32	-0.54	012.01	0
012	Starch (%)	0353	31.76	0.3100	32.63	1.645	0.4713	32	-0.53	012.20	0
012	Starch (%)	0278	31.90	0.6000	32.63	1.645	0.4713	32	-0.45	012.04	0
012	Starch (%)	0407	31.91	1.077	32.63	1.645	0.4713	32	-0.44	012.03	0
012	Starch (%)	2129	32.03	0.1150	32.63	1.645	0.4713	32	-0.37	012.04	0
012	Starch (%)	0045	32.05	0.7000	32.63	1.645	0.4713	32	-0.35	012.01	0
012	Starch (%)	0870	32.13	1.339	32.63	1.645	0.4713	32	-0.30	012.01	0
012	Starch (%)	0889	32.65	0.4300	32.63	1.645	0.4713	32	0.01	012.03	0
012	Starch (%)	0910	33.00	0.0000	32.63	1.645	0.4713	32	0.22	012.00	0
012	Starch (%)	0693	33.09	1.490	32.63	1.645	0.4713	32	0.28	012.01	0
012	Starch (%)	2181	33.21	0.1100	32.63	1.645	0.4713	32	0.35	012.00	0
012	Starch (%)	2192	33.35	0.0300	32.63	1.645	0.4713	32	0.43	012.00	0
012	Starch (%)	0910	33.45	0.9000	32.63	1.645	0.4713	32	0.50	012.01	0
012	Starch (%)	0354	33.48	0.0200	32.63	1.645	0.4713	32	0.51	012.00	0
012	Starch (%)	0683	33.54	0.1100	32.63	1.645	0.4713	32	0.55	012.00	0
012	Starch (%)	2188	33.61	0.1700	32.63	1.645	0.4713	32	0.59	012.00	0
012	Starch (%)	0164	33.95	0.9000	32.63	1.645	0.4713	32	0.80	012.00	0
012	Starch (%)	0948	33.95	0.7720	32.63	1.645	0.4713	32	0.80	012.04	0
012	Starch (%)	2006	34.09	0.8500	32.63	1.645	0.4713	32	0.88	012.00	0
012	Starch (%)	0878	34.09	0.1900	32.63	1.645	0.4713	32	0.88	012.00	0
012	Starch (%)	2196	34.47	0.0000	32.63	1.645	0.4713	32	1.12	012.00	0
012	Starch (%)	0889	34.49	0.0300	32.63	1.645	0.4713	32	1.13	012.11	0
012	Starch (%)	2089	36.53	0.3600	32.63	1.645	0.4713	32	2.37	012.01	0
012	Starch (%)	0297	37.78	0.6400	32.63	1.645	0.4713	32	3.13	012.11	0
013	Fat, Acid Pretreat (%)	0618	3.764	1.509	5.591	0.5441	0.3050	44	-3.36	013.08	0
013	Fat, Acid Pretreat (%)	2073	4.425	0.2900	5.591	0.5441	0.3050	44	-2.14	013.00	0
013	Fat, Acid Pretreat (%)	2146	4.620	0.1200	5.591	0.5441	0.3050	44	-1.79	013.00	0
013	Fat, Acid Pretreat (%)	0083	4.690	0.1000	5.591	0.5441	0.3050	44	-1.66	013.00	0
013	Fat, Acid Pretreat (%)	0036	4.785	0.3236	5.591	0.5441	0.3050	44	-1.48	013.10	0
013	Fat, Acid Pretreat (%)	0939	4.830	0.1800	5.591	0.5441	0.3050	44	-1.40	013.00	0
013	Fat, Acid Pretreat (%)	0638	4.870	0.1320	5.591	0.5441	0.3050	44	-1.33	013.10	0
013	Fat, Acid Pretreat (%)	0541	5.110	0.1600	5.591	0.5441	0.3050	44	-0.88	013.00	0
013	Fat, Acid Pretreat (%)	0208	5.185	0.2100	5.591	0.5441	0.3050	44	-0.75	013.02	0
013	Fat, Acid Pretreat (%)	0968	5.211	0.0240	5.591	0.5441	0.3050	44	-0.70	013.13	0
013	Fat, Acid Pretreat (%)	0297	5.265	0.6700	5.591	0.5441	0.3050	44	-0.60	013.00	0
013	Fat, Acid Pretreat (%)	0921	5.270	0.0600	5.591	0.5441	0.3050	44	-0.59	013.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
013	Fat, Acid Pretreat (%)	0652	5.400	0.2000	5.591	0.5441	0.3050	44	-0.35	013.00	0
013	Fat, Acid Pretreat (%)	2196	5.410	0.0000	5.591	0.5441	0.3050	44	-0.33	013.10	0
013	Fat, Acid Pretreat (%)	0910	5.435	0.1100	5.591	0.5441	0.3050	44	-0.29	013.00	0
013	Fat, Acid Pretreat (%)	0914	5.450	0.1000	5.591	0.5441	0.3050	44	-0.26	013.02	0
013	Fat, Acid Pretreat (%)	2076	5.488	1.012	5.591	0.5441	0.3050	44	-0.19	013.00	0
013	Fat, Acid Pretreat (%)	0229	5.505	0.3500	5.591	0.5441	0.3050	44	-0.16	013.00	0
013	Fat, Acid Pretreat (%)	0853	5.550	0.1000	5.591	0.5441	0.3050	44	-0.08	013.02	0
013	Fat, Acid Pretreat (%)	0948	5.580	0.0200	5.591	0.5441	0.3050	44	-0.02	013.00	0
013	Fat, Acid Pretreat (%)	0353	5.580	0.0200	5.591	0.5441	0.3050	44	-0.02	013.10	0
013	Fat, Acid Pretreat (%)	0098	5.595	0.1500	5.591	0.5441	0.3050	44	0.01	013.02	0
013	Fat, Acid Pretreat (%)	0265	5.595	0.0300	5.591	0.5441	0.3050	44	0.01	013.13	0
013	Fat, Acid Pretreat (%)	0843	5.605	0.5500	5.591	0.5441	0.3050	44	0.03	013.13	0
013	Fat, Acid Pretreat (%)	0309	5.674	0.2364	5.591	0.5441	0.3050	44	0.15	013.00	0
013	Fat, Acid Pretreat (%)	0682	5.760	0.0000	5.591	0.5441	0.3050	44	0.31	013.02	0
013	Fat, Acid Pretreat (%)	0148	5.770	0.1800	5.591	0.5441	0.3050	44	0.33	013.02	0
013	Fat, Acid Pretreat (%)	0643	5.770	0.6600	5.591	0.5441	0.3050	44	0.33	013.02	0
013	Fat, Acid Pretreat (%)	0354	5.790	0.1400	5.591	0.5441	0.3050	44	0.37	013.02	0
013	Fat, Acid Pretreat (%)	0884	5.800	0.4000	5.591	0.5441	0.3050	44	0.38	013.00	0
013	Fat, Acid Pretreat (%)	0100	5.800	0.3600	5.591	0.5441	0.3050	44	0.38	013.02	0
013	Fat, Acid Pretreat (%)	2304	5.820	0.1400	5.591	0.5441	0.3050	44	0.42	013.00	0
013	Fat, Acid Pretreat (%)	2150	5.895	0.0700	5.591	0.5441	0.3050	44	0.56	013.00	0
013	Fat, Acid Pretreat (%)	0171	5.990	0.2800	5.591	0.5441	0.3050	44	0.73	013.02	0
013	Fat, Acid Pretreat (%)	0581	6.040	0.1200	5.591	0.5441	0.3050	44	0.82	013.13	0
013	Fat, Acid Pretreat (%)	0870	6.070	1.318	5.591	0.5441	0.3050	44	0.88	013.02	0
013	Fat, Acid Pretreat (%)	0139	6.100	0.1200	5.591	0.5441	0.3050	44	0.93	013.00	0
013	Fat, Acid Pretreat (%)	0202	6.105	0.2500	5.591	0.5441	0.3050	44	0.94	013.00	0
013	Fat, Acid Pretreat (%)	0675	6.120	0.0200	5.591	0.5441	0.3050	44	0.97	013.02	0
013	Fat, Acid Pretreat (%)	0164	6.165	0.0900	5.591	0.5441	0.3050	44	1.05	013.02	0
013	Fat, Acid Pretreat (%)	0407	6.330	0.0420	5.591	0.5441	0.3050	44	1.36	013.02	0
013	Fat, Acid Pretreat (%)	2259	6.387	0.0690	5.591	0.5441	0.3050	44	1.46	013.02	0
013	Fat, Acid Pretreat (%)	0676	6.805	1.050	5.591	0.5441	0.3050	44	2.23	013.02	0
013	Fat, Acid Pretreat (%)	0407	7.000	1.453	5.591	0.5441	0.3050	44	2.59	013.13	0
013	Fat, Acid Pretreat (%)	0618	6.934	2.049	5.591	0.5441	0.3050	44	2.47	013.00	1
015	Aluminum (ppm)	0037	51.50	1.000	187.2	30.29	7.617	18	-4.48	015.42	0
015	Aluminum (ppm)	0560	141.5	27.20	187.2	30.29	7.617	18	-1.51	015.52	0
015	Aluminum (ppm)	2129	157.7	7.200	187.2	30.29	7.617	18	-0.97	015.42	0
015	Aluminum (ppm)	0407	164.5	1.492	187.2	30.29	7.617	18	-0.75	015.41	0
015	Aluminum (ppm)	0171	165.9	0.0000	187.2	30.29	7.617	18	-0.70	015.41	0
015	Aluminum (ppm)	0918	177.1	8.379	187.2	30.29	7.617	18	-0.33	015.43	0
015	Aluminum (ppm)	0510	182.0	4.000	187.2	30.29	7.617	18	-0.17	015.43	0
015	Aluminum (ppm)	0164	184.5	7.000	187.2	30.29	7.617	18	-0.09	015.41	0
015	Aluminum (ppm)	0169	184.5	1.000	187.2	30.29	7.617	18	-0.09	015.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
015	Aluminum (ppm)	0345	188.3	1.500	187.2	30.29	7.617	18	0.04	015.43	0
015	Aluminum (ppm)	0520	188.5	3.000	187.2	30.29	7.617	18	0.04	015.41	0
015	Aluminum (ppm)	0297	191.0	4.000	187.2	30.29	7.617	18	0.13	015.43	0
015	Aluminum (ppm)	0425	206.1	19.50	187.2	30.29	7.617	18	0.62	015.43	0
015	Aluminum (ppm)	0353	206.5	13.00	187.2	30.29	7.617	18	0.64	015.43	0
015	Aluminum (ppm)	0407	208.8	10.84	187.2	30.29	7.617	18	0.71	015.43	0
015	Aluminum (ppm)	0553	218.0	2.000	187.2	30.29	7.617	18	1.02	015.53	0
015	Aluminum (ppm)	0939	229.5	19.00	187.2	30.29	7.617	18	1.40	015.53	0
015	Aluminum (ppm)	2302	266.5	7.000	187.2	30.29	7.617	18	2.62	015.99	0
017	Boron (ppm)	2053	10.28	0.2700	11.62	0.9270	0.9967	19	-1.45	017.42	0
017	Boron (ppm)	0096	10.49	1.711	11.62	0.9270	0.9967	19	-1.22	017.52	0
017	Boron (ppm)	0345	10.50	0.6000	11.62	0.9270	0.9967	19	-1.20	017.43	0
017	Boron (ppm)	0265	10.65	0.5000	11.62	0.9270	0.9967	19	-1.04	017.42	0
017	Boron (ppm)	0693	10.83	0.6230	11.62	0.9270	0.9967	19	-0.84	017.42	0
017	Boron (ppm)	0560	11.38	3.060	11.62	0.9270	0.9967	19	-0.26	017.52	0
017	Boron (ppm)	2129	11.50	0.1000	11.62	0.9270	0.9967	19	-0.13	017.42	0
017	Boron (ppm)	0407	11.51	0.9282	11.62	0.9270	0.9967	19	-0.12	017.41	0
017	Boron (ppm)	0083	11.60	0.2000	11.62	0.9270	0.9967	19	-0.02	017.43	0
017	Boron (ppm)	0037	11.60	0.4000	11.62	0.9270	0.9967	19	-0.02	017.42	0
017	Boron (ppm)	0358	11.66	0.6700	11.62	0.9270	0.9967	19	0.04	017.41	0
017	Boron (ppm)	0045	11.70	0.6000	11.62	0.9270	0.9967	19	0.09	017.42	0
017	Boron (ppm)	0226	11.73	1.010	11.62	0.9270	0.9967	19	0.12	017.41	0
017	Boron (ppm)	0294	12.15	0.3000	11.62	0.9270	0.9967	19	0.58	017.42	0
017	Boron (ppm)	0407	12.16	0.1948	11.62	0.9270	0.9967	19	0.58	017.43	0
017	Boron (ppm)	0918	12.27	0.1300	11.62	0.9270	0.9967	19	0.70	017.43	0
017	Boron (ppm)	0229	12.71	3.640	11.62	0.9270	0.9967	19	1.18	017.41	0
017	Boron (ppm)	0353	14.00	4.000	11.62	0.9270	0.9967	19	2.57	017.43	0
017	Boron (ppm)	0510	14.00	0.0000	11.62	0.9270	0.9967	19	2.57	017.43	0
017	Boron (ppm)	0297	0.0000	0.0000	11.62	0.9270	0.9967	19		017.43	4
019	Calcium (%)	0096	1.338	0.0950	1.543	0.0795	0.0367	115	-2.58	019.42	0
019	Calcium (%)	0154	1.343	0.0715	1.543	0.0795	0.0367	115	-2.51	019.52	0
019	Calcium (%)	0018	1.345	0.0100	1.543	0.0795	0.0367	115	-2.49	019.41	0
019	Calcium (%)	0089	1.350	0.0000	1.543	0.0795	0.0367	115	-2.42	019.43	0
019	Calcium (%)	2188	1.375	0.0100	1.543	0.0795	0.0367	115	-2.11	019.31	0
019	Calcium (%)	2129	1.405	0.0110	1.543	0.0795	0.0367	115	-1.74	019.42	0
019	Calcium (%)	0848	1.405	0.0300	1.543	0.0795	0.0367	115	-1.73	019.41	0
019	Calcium (%)	2192	1.425	0.0500	1.543	0.0795	0.0367	115	-1.48	019.43	0
019	Calcium (%)	0964	1.430	0.0614	1.543	0.0795	0.0367	115	-1.43	019.43	0
019	Calcium (%)	2188	1.435	0.0100	1.543	0.0795	0.0367	115	-1.36	019.08	0
019	Calcium (%)	0175	1.435	0.0300	1.543	0.0795	0.0367	115	-1.36	019.31	0
019	Calcium (%)	0520	1.440	0.0000	1.543	0.0795	0.0367	115	-1.29	019.41	0
019	Calcium (%)	0357	1.453	0.0053	1.543	0.0795	0.0367	115	-1.13	019.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
019	Calcium (%)	0036	1.455	0.0155	1.543	0.0795	0.0367	115	-1.11	019.43	0
019	Calcium (%)	0148	1.455	0.1500	1.543	0.0795	0.0367	115	-1.10	019.41	0
019	Calcium (%)	0910	1.455	0.1500	1.543	0.0795	0.0367	115	-1.10	019.41	0
019	Calcium (%)	2006	1.457	0.0250	1.543	0.0795	0.0367	115	-1.09	019.00	0
019	Calcium (%)	2146	1.460	0.1600	1.543	0.0795	0.0367	115	-1.04	019.31	0
019	Calcium (%)	0199	1.460	0.0400	1.543	0.0795	0.0367	115	-1.04	019.53	0
019	Calcium (%)	0674	1.470	0.0000	1.543	0.0795	0.0367	115	-0.92	019.31	0
019	Calcium (%)	0590	1.470	0.1400	1.543	0.0795	0.0367	115	-0.92	019.99	0
019	Calcium (%)	2259	1.473	0.0080	1.543	0.0795	0.0367	115	-0.88	019.41	0
019	Calcium (%)	0884	1.475	0.0500	1.543	0.0795	0.0367	115	-0.85	019.31	0
019	Calcium (%)	0169	1.475	0.0100	1.543	0.0795	0.0367	115	-0.85	019.32	0
019	Calcium (%)	0142	1.475	0.0606	1.543	0.0795	0.0367	115	-0.85	019.31	0
019	Calcium (%)	0968	1.477	0.0040	1.543	0.0795	0.0367	115	-0.83	019.43	0
019	Calcium (%)	0941	1.488	0.0000	1.543	0.0795	0.0367	115	-0.69	019.43	0
019	Calcium (%)	0298	1.490	0.0200	1.543	0.0795	0.0367	115	-0.66	019.41	0
019	Calcium (%)	2144	1.495	0.0100	1.543	0.0795	0.0367	115	-0.60	019.00	0
019	Calcium (%)	0144	1.495	0.0100	1.543	0.0795	0.0367	115	-0.60	019.41	0
019	Calcium (%)	0918	1.499	0.0560	1.543	0.0795	0.0367	115	-0.55	019.43	0
019	Calcium (%)	0019	1.500	0.1200	1.543	0.0795	0.0367	115	-0.54	019.41	0
019	Calcium (%)	0675	1.505	0.0500	1.543	0.0795	0.0367	115	-0.48	019.43	0
019	Calcium (%)	0208	1.508	0.0110	1.543	0.0795	0.0367	115	-0.44	019.31	0
019	Calcium (%)	0948	1.510	0.0020	1.543	0.0795	0.0367	115	-0.41	019.31	0
019	Calcium (%)	0139	1.510	0.0000	1.543	0.0795	0.0367	115	-0.41	019.41	0
019	Calcium (%)	0921	1.511	0.0190	1.543	0.0795	0.0367	115	-0.41	019.31	0
019	Calcium (%)	0878	1.515	0.0300	1.543	0.0795	0.0367	115	-0.35	019.00	0
019	Calcium (%)	0878	1.515	0.0100	1.543	0.0795	0.0367	115	-0.35	019.41	0
019	Calcium (%)	2053	1.515	0.0100	1.543	0.0795	0.0367	115	-0.35	019.42	0
019	Calcium (%)	0164	1.515	0.0100	1.543	0.0795	0.0367	115	-0.35	019.44	0
019	Calcium (%)	0939	1.515	0.0100	1.543	0.0795	0.0367	115	-0.35	019.53	0
019	Calcium (%)	0263	1.516	0.0030	1.543	0.0795	0.0367	115	-0.34	019.42	0
019	Calcium (%)	2196	1.520	0.0000	1.543	0.0795	0.0367	115	-0.29	019.31	0
019	Calcium (%)	2012	1.520	0.0000	1.543	0.0795	0.0367	115	-0.29	019.41	0
019	Calcium (%)	0510	1.520	0.0200	1.543	0.0795	0.0367	115	-0.29	019.43	0
019	Calcium (%)	0868	1.520	0.0200	1.543	0.0795	0.0367	115	-0.29	019.99	0
019	Calcium (%)	0345	1.525	0.0100	1.543	0.0795	0.0367	115	-0.22	019.43	0
019	Calcium (%)	0572	1.525	0.0900	1.543	0.0795	0.0367	115	-0.22	019.53	0
019	Calcium (%)	2196	1.526	0.0000	1.543	0.0795	0.0367	115	-0.21	019.08	0
019	Calcium (%)	0676	1.530	0.0000	1.543	0.0795	0.0367	115	-0.16	019.99	0
019	Calcium (%)	0407	1.533	0.0009	1.543	0.0795	0.0367	115	-0.12	019.41	0
019	Calcium (%)	2062	1.535	0.0280	1.543	0.0795	0.0367	115	-0.10	019.00	0
019	Calcium (%)	0229	1.535	0.0100	1.543	0.0795	0.0367	115	-0.10	019.41	0
019	Calcium (%)	0278	1.535	0.0900	1.543	0.0795	0.0367	115	-0.10	019.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
019	Calcium (%)	0693	1.536	0.0500	1.543	0.0795	0.0367	115	-0.09	019.42	0
019	Calcium (%)	0881	1.539	0.0970	1.543	0.0795	0.0367	115	-0.05	019.00	0
019	Calcium (%)	0074	1.540	0.0000	1.543	0.0795	0.0367	115	-0.04	019.41	0
019	Calcium (%)	0297	1.540	0.0600	1.543	0.0795	0.0367	115	-0.04	019.43	0
019	Calcium (%)	0590	1.545	0.0500	1.543	0.0795	0.0367	115	0.03	019.08	0
019	Calcium (%)	0035	1.545	0.0700	1.543	0.0795	0.0367	115	0.03	019.42	0
019	Calcium (%)	0083	1.545	0.0100	1.543	0.0795	0.0367	115	0.03	019.43	0
019	Calcium (%)	2146	1.545	0.0300	1.543	0.0795	0.0367	115	0.03	019.51	0
019	Calcium (%)	2181	1.550	0.0000	1.543	0.0795	0.0367	115	0.09	019.00	0
019	Calcium (%)	0656	1.550	0.0000	1.543	0.0795	0.0367	115	0.09	019.31	0
019	Calcium (%)	0425	1.550	0.1000	1.543	0.0795	0.0367	115	0.09	019.43	0
019	Calcium (%)	2089	1.550	0.0000	1.543	0.0795	0.0367	115	0.09	019.43	0
019	Calcium (%)	2295	1.555	0.0100	1.543	0.0795	0.0367	115	0.15	019.00	0
019	Calcium (%)	0868	1.560	0.0400	1.543	0.0795	0.0367	115	0.22	019.31	0
019	Calcium (%)	0529	1.565	0.0700	1.543	0.0795	0.0367	115	0.28	019.31	0
019	Calcium (%)	0294	1.565	0.0100	1.543	0.0795	0.0367	115	0.28	019.42	0
019	Calcium (%)	2109	1.569	0.0050	1.543	0.0795	0.0367	115	0.32	019.00	0
019	Calcium (%)	0358	1.570	0.0800	1.543	0.0795	0.0367	115	0.34	019.41	0
019	Calcium (%)	0010	1.570	0.1000	1.543	0.0795	0.0367	115	0.34	019.43	0
019	Calcium (%)	0505	1.570	0.0200	1.543	0.0795	0.0367	115	0.34	019.43	0
019	Calcium (%)	0687	1.575	0.1500	1.543	0.0795	0.0367	115	0.40	019.31	0
019	Calcium (%)	0123	1.575	0.0100	1.543	0.0795	0.0367	115	0.40	019.41	0
019	Calcium (%)	0171	1.575	0.0100	1.543	0.0795	0.0367	115	0.40	019.41	0
019	Calcium (%)	0202	1.575	0.0700	1.543	0.0795	0.0367	115	0.40	019.43	0
019	Calcium (%)	0009	1.579	0.0530	1.543	0.0795	0.0367	115	0.45	019.42	0
019	Calcium (%)	0037	1.580	0.0130	1.543	0.0795	0.0367	115	0.46	019.42	0
019	Calcium (%)	0354	1.585	0.0100	1.543	0.0795	0.0367	115	0.53	019.41	0
019	Calcium (%)	0035	1.585	0.0100	1.543	0.0795	0.0367	115	0.53	019.43	0
019	Calcium (%)	2234	1.585	0.0300	1.543	0.0795	0.0367	115	0.53	019.43	0
019	Calcium (%)	0870	1.588	0.0200	1.543	0.0795	0.0367	115	0.57	019.42	0
019	Calcium (%)	2302	1.590	0.0200	1.543	0.0795	0.0367	115	0.59	019.00	0
019	Calcium (%)	0939	1.590	0.0200	1.543	0.0795	0.0367	115	0.59	019.31	0
019	Calcium (%)	2006	1.595	0.0050	1.543	0.0795	0.0367	115	0.65	019.09	0
019	Calcium (%)	0683	1.595	0.0100	1.543	0.0795	0.0367	115	0.66	019.00	0
019	Calcium (%)	0265	1.600	0.1400	1.543	0.0795	0.0367	115	0.72	019.42	0
019	Calcium (%)	0208	1.602	0.0260	1.543	0.0795	0.0367	115	0.74	019.41	0
019	Calcium (%)	0066	1.603	0.0260	1.543	0.0795	0.0367	115	0.76	019.41	0
019	Calcium (%)	0100	1.605	0.0700	1.543	0.0795	0.0367	115	0.78	019.99	0
019	Calcium (%)	0692	1.610	0.0600	1.543	0.0795	0.0367	115	0.84	019.42	0
019	Calcium (%)	0096	1.615	0.0430	1.543	0.0795	0.0367	115	0.90	019.52	0
019	Calcium (%)	0553	1.620	0.0200	1.543	0.0795	0.0367	115	0.97	019.53	0
019	Calcium (%)	0885	1.624	0.0048	1.543	0.0795	0.0367	115	1.02	019.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	2302	1.625	0.0300	1.543	0.0795	0.0367	115	1.03	019.99	0
019	Calcium (%)	0226	1.630	0.0200	1.543	0.0795	0.0367	115	1.10	019.41	0
019	Calcium (%)	0682	1.630	0.0000	1.543	0.0795	0.0367	115	1.10	019.43	0
019	Calcium (%)	0536	1.635	0.0100	1.543	0.0795	0.0367	115	1.16	019.31	0
019	Calcium (%)	0042	1.650	0.0800	1.543	0.0795	0.0367	115	1.35	019.42	0
019	Calcium (%)	0001	1.654	0.0370	1.543	0.0795	0.0367	115	1.39	019.31	0
019	Calcium (%)	0598	1.654	0.0190	1.543	0.0795	0.0367	115	1.39	019.41	0
019	Calcium (%)	0036	1.658	0.0032	1.543	0.0795	0.0367	115	1.45	019.03	0
019	Calcium (%)	0032	1.659	0.0020	1.543	0.0795	0.0367	115	1.46	019.43	0
019	Calcium (%)	0098	1.665	0.0300	1.543	0.0795	0.0367	115	1.54	019.41	0
019	Calcium (%)	0045	1.670	0.0000	1.543	0.0795	0.0367	115	1.60	019.42	0
019	Calcium (%)	0726	1.670	0.1200	1.543	0.0795	0.0367	115	1.60	019.42	0
019	Calcium (%)	0003	1.675	0.0500	1.543	0.0795	0.0367	115	1.66	019.41	0
019	Calcium (%)	0612	1.675	0.1100	1.543	0.0795	0.0367	115	1.66	019.33	0
019	Calcium (%)	0407	1.678	0.0291	1.543	0.0795	0.0367	115	1.70	019.43	0
019	Calcium (%)	0190	1.710	0.0200	1.543	0.0795	0.0367	115	2.10	019.42	0
019	Calcium (%)	0098	1.720	0.1400	1.543	0.0795	0.0367	115	2.23	019.43	0
019	Calcium (%)	0353	1.730	0.0600	1.543	0.0795	0.0367	115	2.35	019.43	0
019	Calcium (%)	0560	1.527	0.2010	1.543	0.0795	0.0367	115	-0.21	019.52	1
021	Cobalt (ppm)	0171	4.225	0.0100	5.540	0.6899	0.2701	25	-1.91	021.41	0
021	Cobalt (ppm)	0693	4.718	0.7580	5.540	0.6899	0.2701	25	-1.19	021.42	0
021	Cobalt (ppm)	0199	4.785	0.2700	5.540	0.6899	0.2701	25	-1.09	021.53	0
021	Cobalt (ppm)	0407	4.812	0.1414	5.540	0.6899	0.2701	25	-1.06	021.41	0
021	Cobalt (ppm)	0345	4.995	0.1100	5.540	0.6899	0.2701	25	-0.79	021.43	0
021	Cobalt (ppm)	0186	5.035	0.0100	5.540	0.6899	0.2701	25	-0.73	021.52	0
021	Cobalt (ppm)	0910	5.035	0.0300	5.540	0.6899	0.2701	25	-0.73	021.52	0
021	Cobalt (ppm)	0560	5.110	0.6000	5.540	0.6899	0.2701	25	-0.62	021.52	0
021	Cobalt (ppm)	2053	5.190	0.3200	5.540	0.6899	0.2701	25	-0.51	021.42	0
021	Cobalt (ppm)	0918	5.294	0.0418	5.540	0.6899	0.2701	25	-0.36	021.53	0
021	Cobalt (ppm)	2234	5.360	0.6600	5.540	0.6899	0.2701	25	-0.26	021.43	0
021	Cobalt (ppm)	0169	5.380	0.1400	5.540	0.6899	0.2701	25	-0.23	021.43	0
021	Cobalt (ppm)	0510	5.520	0.0600	5.540	0.6899	0.2701	25	-0.03	021.43	0
021	Cobalt (ppm)	0066	5.544	0.2660	5.540	0.6899	0.2701	25	0.01	021.41	0
021	Cobalt (ppm)	0407	5.551	0.0962	5.540	0.6899	0.2701	25	0.02	021.43	0
021	Cobalt (ppm)	0096	5.642	1.018	5.540	0.6899	0.2701	25	0.15	021.52	0
021	Cobalt (ppm)	0939	5.725	0.0900	5.540	0.6899	0.2701	25	0.27	021.53	0
021	Cobalt (ppm)	0164	5.950	0.3000	5.540	0.6899	0.2701	25	0.59	021.31	0
021	Cobalt (ppm)	0148	6.085	0.2300	5.540	0.6899	0.2701	25	0.79	021.41	0
021	Cobalt (ppm)	0941	6.150	0.0000	5.540	0.6899	0.2701	25	0.88	021.43	0
021	Cobalt (ppm)	0045	6.325	0.6500	5.540	0.6899	0.2701	25	1.14	021.42	0
021	Cobalt (ppm)	0939	6.350	0.2400	5.540	0.6899	0.2701	25	1.17	021.31	0
021	Cobalt (ppm)	0682	6.400	0.0000	5.540	0.6899	0.2701	25	1.25	021.43	0



Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
021	Cobalt (ppm)	0553	6.455	0.1500	5.540	0.6899	0.2701	25	1.33	021.53	0
021	Cobalt (ppm)	0353	6.859	0.5607	5.540	0.6899	0.2701	25	1.91	021.43	0
022	Copper (ppm)	0536	61.55	0.6000	76.85	6.521	3.239	84	-2.35	022.31	0
022	Copper (ppm)	0003	62.00	4.000	76.85	6.521	3.239	84	-2.28	022.41	0
022	Copper (ppm)	0590	62.05	0.5000	76.85	6.521	3.239	84	-2.27	022.99	0
022	Copper (ppm)	0598	62.23	2.140	76.85	6.521	3.239	84	-2.24	022.41	0
022	Copper (ppm)	2012	62.34	0.2200	76.85	6.521	3.239	84	-2.22	022.41	0
022	Copper (ppm)	0354	63.51	0.1000	76.85	6.521	3.239	84	-2.05	022.41	0
022	Copper (ppm)	2146	65.93	0.3000	76.85	6.521	3.239	84	-1.67	022.31	0
022	Copper (ppm)	2062	66.68	6.309	76.85	6.521	3.239	84	-1.56	022.31	0
022	Copper (ppm)	0001	67.12	4.980	76.85	6.521	3.239	84	-1.49	022.31	0
022	Copper (ppm)	0948	69.09	0.7270	76.85	6.521	3.239	84	-1.19	022.33	0
022	Copper (ppm)	0148	69.15	3.700	76.85	6.521	3.239	84	-1.18	022.41	0
022	Copper (ppm)	0019	69.36	4.810	76.85	6.521	3.239	84	-1.15	022.41	0
022	Copper (ppm)	0407	69.38	1.481	76.85	6.521	3.239	84	-1.14	022.41	0
022	Copper (ppm)	2302	69.50	13.00	76.85	6.521	3.239	84	-1.13	022.99	0
022	Copper (ppm)	0868	69.98	0.1200	76.85	6.521	3.239	84	-1.05	022.31	0
022	Copper (ppm)	0164	70.00	2.000	76.85	6.521	3.239	84	-1.05	022.44	0
022	Copper (ppm)	0171	70.95	2.300	76.85	6.521	3.239	84	-0.90	022.41	0
022	Copper (ppm)	0529	71.00	1.000	76.85	6.521	3.239	84	-0.90	022.31	0
022	Copper (ppm)	0910	71.00	2.000	76.85	6.521	3.239	84	-0.90	022.41	0
022	Copper (ppm)	0921	71.80	8.600	76.85	6.521	3.239	84	-0.77	022.31	0
022	Copper (ppm)	0848	72.44	0.1700	76.85	6.521	3.239	84	-0.68	022.41	0
022	Copper (ppm)	0918	72.51	2.940	76.85	6.521	3.239	84	-0.67	022.43	0
022	Copper (ppm)	0226	72.69	4.290	76.85	6.521	3.239	84	-0.64	022.41	0
022	Copper (ppm)	0294	73.18	1.050	76.85	6.521	3.239	84	-0.56	022.42	0
022	Copper (ppm)	0520	73.50	5.000	76.85	6.521	3.239	84	-0.51	022.41	0
022	Copper (ppm)	0278	73.75	6.500	76.85	6.521	3.239	84	-0.48	022.42	0
022	Copper (ppm)	0098	74.46	9.320	76.85	6.521	3.239	84	-0.37	022.41	0
022	Copper (ppm)	0074	74.50	11.00	76.85	6.521	3.239	84	-0.36	022.41	0
022	Copper (ppm)	0560	74.55	11.90	76.85	6.521	3.239	84	-0.35	022.52	0
022	Copper (ppm)	2188	74.91	1.700	76.85	6.521	3.239	84	-0.30	022.31	0
022	Copper (ppm)	0692	75.00	0.0000	76.85	6.521	3.239	84	-0.28	022.42	0
022	Copper (ppm)	0358	75.03	5.660	76.85	6.521	3.239	84	-0.28	022.41	0
022	Copper (ppm)	0553	75.05	1.900	76.85	6.521	3.239	84	-0.28	022.53	0
022	Copper (ppm)	0884	75.15	0.5000	76.85	6.521	3.239	84	-0.26	022.31	0
022	Copper (ppm)	0202	75.83	1.650	76.85	6.521	3.239	84	-0.16	022.43	0
022	Copper (ppm)	0878	76.50	5.000	76.85	6.521	3.239	84	-0.05	022.41	0
022	Copper (ppm)	0066	77.03	0.7100	76.85	6.521	3.239	84	0.03	022.44	0
022	Copper (ppm)	0263	77.32	1.070	76.85	6.521	3.239	84	0.07	022.42	0
022	Copper (ppm)	0345	77.32	1.570	76.85	6.521	3.239	84	0.07	022.43	0
022	Copper (ppm)	2196	77.32	0.0000	76.85	6.521	3.239	84	0.07	022.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
022	Copper (ppm)	0939	77.42	0.6900	76.85	6.521	3.239	84	0.09	022.53	0
022	Copper (ppm)	0941	77.50	0.0000	76.85	6.521	3.239	84	0.10	022.43	0
022	Copper (ppm)	0939	77.53	1.020	76.85	6.521	3.239	84	0.10	022.31	0
022	Copper (ppm)	0229	77.82	6.740	76.85	6.521	3.239	84	0.15	022.41	0
022	Copper (ppm)	0175	78.00	0.0000	76.85	6.521	3.239	84	0.18	022.31	0
022	Copper (ppm)	0100	78.00	0.0000	76.85	6.521	3.239	84	0.18	022.99	0
022	Copper (ppm)	0169	78.10	2.000	76.85	6.521	3.239	84	0.19	022.43	0
022	Copper (ppm)	0208	78.43	2.780	76.85	6.521	3.239	84	0.24	022.41	0
022	Copper (ppm)	0505	78.49	4.690	76.85	6.521	3.239	84	0.25	022.43	0
022	Copper (ppm)	0083	78.50	1.000	76.85	6.521	3.239	84	0.25	022.43	0
022	Copper (ppm)	0032	78.98	0.4200	76.85	6.521	3.239	84	0.33	022.43	0
022	Copper (ppm)	0425	79.35	13.50	76.85	6.521	3.239	84	0.38	022.43	0
022	Copper (ppm)	0186	79.50	7.000	76.85	6.521	3.239	84	0.41	022.42	0
022	Copper (ppm)	0035	79.50	1.000	76.85	6.521	3.239	84	0.41	022.43	0
022	Copper (ppm)	0186	79.60	6.200	76.85	6.521	3.239	84	0.42	022.52	0
022	Copper (ppm)	0096	79.71	9.599	76.85	6.521	3.239	84	0.44	022.42	0
022	Copper (ppm)	0357	79.86	1.045	76.85	6.521	3.239	84	0.46	022.42	0
022	Copper (ppm)	0656	80.11	0.0000	76.85	6.521	3.239	84	0.50	022.31	0
022	Copper (ppm)	0265	80.40	7.600	76.85	6.521	3.239	84	0.54	022.42	0
022	Copper (ppm)	0009	80.50	3.000	76.85	6.521	3.239	84	0.56	022.42	0
022	Copper (ppm)	2089	80.54	0.3000	76.85	6.521	3.239	84	0.57	022.43	0
022	Copper (ppm)	0045	80.65	1.100	76.85	6.521	3.239	84	0.58	022.42	0
022	Copper (ppm)	2129	81.41	0.6100	76.85	6.521	3.239	84	0.70	022.42	0
022	Copper (ppm)	0682	81.90	0.0000	76.85	6.521	3.239	84	0.77	022.43	0
022	Copper (ppm)	0870	81.91	2.150	76.85	6.521	3.239	84	0.78	022.42	0
022	Copper (ppm)	0297	82.00	10.00	76.85	6.521	3.239	84	0.79	022.43	0
022	Copper (ppm)	0968	82.00	2.000	76.85	6.521	3.239	84	0.79	022.43	0
022	Copper (ppm)	0572	82.35	4.700	76.85	6.521	3.239	84	0.84	022.53	0
022	Copper (ppm)	0726	82.64	1.660	76.85	6.521	3.239	84	0.89	022.42	0
022	Copper (ppm)	0675	83.01	3.020	76.85	6.521	3.239	84	0.94	022.43	0
022	Copper (ppm)	0035	83.50	7.000	76.85	6.521	3.239	84	1.02	022.42	0
022	Copper (ppm)	0510	83.50	1.000	76.85	6.521	3.239	84	1.02	022.43	0
022	Copper (ppm)	0096	83.54	11.56	76.85	6.521	3.239	84	1.03	022.52	0
022	Copper (ppm)	0098	83.81	0.1400	76.85	6.521	3.239	84	1.07	022.43	0
022	Copper (ppm)	0010	83.95	3.900	76.85	6.521	3.239	84	1.09	022.33	0
022	Copper (ppm)	0208	84.05	2.300	76.85	6.521	3.239	84	1.10	022.31	0
022	Copper (ppm)	0190	84.46	0.3700	76.85	6.521	3.239	84	1.17	022.42	0
022	Copper (ppm)	0693	84.53	0.1340	76.85	6.521	3.239	84	1.18	022.42	0
022	Copper (ppm)	0407	84.61	0.3271	76.85	6.521	3.239	84	1.19	022.43	0
022	Copper (ppm)	2053	85.45	0.1000	76.85	6.521	3.239	84	1.32	022.42	0
022	Copper (ppm)	0199	85.46	7.090	76.85	6.521	3.239	84	1.32	022.53	0
022	Copper (ppm)	0042	88.80	5.600	76.85	6.521	3.239	84	1.83	022.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
022	Copper (ppm)	2234	91.20	2.090	76.85	6.521	3.239	84	2.20	022.43	0
022	Copper (ppm)	0037	113.4	5.800	76.85	6.521	3.239	84	5.61	022.42	0
022	Copper (ppm)	0353	96.00	20.00	76.85	6.521	3.239	84	2.94	022.43	1
025	Iron (ppm)	0037	79.70	1.600	459.9	45.52	14.33	71	-8.35	025.42	0
025	Iron (ppm)	0278	130.0	44.00	459.9	45.52	14.33	71	-7.25	025.42	0
025	Iron (ppm)	0726	260.5	8.600	459.9	45.52	14.33	71	-4.38	025.42	0
025	Iron (ppm)	0265	286.0	16.00	459.9	45.52	14.33	71	-3.82	025.42	0
025	Iron (ppm)	2053	321.0	40.00	459.9	45.52	14.33	71	-3.05	025.42	0
025	Iron (ppm)	0199	337.3	7.490	459.9	45.52	14.33	71	-2.69	025.53	0
025	Iron (ppm)	0598	361.7	0.0000	459.9	45.52	14.33	71	-2.16	025.41	0
025	Iron (ppm)	0921	362.5	35.10	459.9	45.52	14.33	71	-2.14	025.31	0
025	Iron (ppm)	0190	393.1	0.8400	459.9	45.52	14.33	71	-1.47	025.42	0
025	Iron (ppm)	0294	393.5	1.000	459.9	45.52	14.33	71	-1.46	025.42	0
025	Iron (ppm)	0096	413.0	45.93	459.9	45.52	14.33	71	-1.03	025.42	0
025	Iron (ppm)	0675	424.6	3.530	459.9	45.52	14.33	71	-0.78	025.43	0
025	Iron (ppm)	0003	425.0	84.00	459.9	45.52	14.33	71	-0.77	025.41	0
025	Iron (ppm)	0042	426.5	95.00	459.9	45.52	14.33	71	-0.73	025.42	0
025	Iron (ppm)	0354	426.5	5.270	459.9	45.52	14.33	71	-0.73	025.41	0
025	Iron (ppm)	2129	429.0	12.70	459.9	45.52	14.33	71	-0.68	025.42	0
025	Iron (ppm)	0263	430.0	7.630	459.9	45.52	14.33	71	-0.66	025.42	0
025	Iron (ppm)	2302	436.5	15.00	459.9	45.52	14.33	71	-0.51	025.99	0
025	Iron (ppm)	0226	437.8	3.800	459.9	45.52	14.33	71	-0.48	025.41	0
025	Iron (ppm)	0656	438.1	0.0000	459.9	45.52	14.33	71	-0.48	025.31	0
025	Iron (ppm)	0692	439.0	30.00	459.9	45.52	14.33	71	-0.46	025.42	0
025	Iron (ppm)	0345	439.0	12.00	459.9	45.52	14.33	71	-0.46	025.43	0
025	Iron (ppm)	0520	441.0	6.000	459.9	45.52	14.33	71	-0.41	025.41	0
025	Iron (ppm)	0529	442.7	0.6000	459.9	45.52	14.33	71	-0.38	025.31	0
025	Iron (ppm)	0870	443.4	28.30	459.9	45.52	14.33	71	-0.36	025.42	0
025	Iron (ppm)	2146	444.2	11.98	459.9	45.52	14.33	71	-0.34	025.31	0
025	Iron (ppm)	2196	450.5	0.0000	459.9	45.52	14.33	71	-0.21	025.31	0
025	Iron (ppm)	0910	452.5	7.000	459.9	45.52	14.33	71	-0.16	025.41	0
025	Iron (ppm)	0169	455.5	1.000	459.9	45.52	14.33	71	-0.10	025.43	0
025	Iron (ppm)	0968	456.0	4.000	459.9	45.52	14.33	71	-0.09	025.43	0
025	Iron (ppm)	2012	459.6	5.460	459.9	45.52	14.33	71	-0.01	025.41	0
025	Iron (ppm)	0407	459.9	13.82	459.9	45.52	14.33	71	0.00	025.41	0
025	Iron (ppm)	0148	460.0	18.00	459.9	45.52	14.33	71	0.00	025.41	0
025	Iron (ppm)	0358	460.5	23.06	459.9	45.52	14.33	71	0.01	025.41	0
025	Iron (ppm)	0164	462.5	1.000	459.9	45.52	14.33	71	0.06	025.41	0
025	Iron (ppm)	0553	462.5	1.000	459.9	45.52	14.33	71	0.06	025.53	0
025	Iron (ppm)	0297	464.0	28.00	459.9	45.52	14.33	71	0.09	025.43	0
025	Iron (ppm)	0175	465.0	26.00	459.9	45.52	14.33	71	0.11	025.31	0
025	Iron (ppm)	0171	465.0	0.0000	459.9	45.52	14.33	71	0.11	025.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
025	Iron (ppm)	0693	467.3	20.91	459.9	45.52	14.33	71	0.16	025.42	0
025	Iron (ppm)	0074	468.5	15.00	459.9	45.52	14.33	71	0.19	025.41	0
025	Iron (ppm)	0096	468.9	22.50	459.9	45.52	14.33	71	0.20	025.52	0
025	Iron (ppm)	0868	469.0	10.00	459.9	45.52	14.33	71	0.20	025.31	0
025	Iron (ppm)	0425	469.2	40.00	459.9	45.52	14.33	71	0.20	025.43	0
025	Iron (ppm)	0019	470.5	11.80	459.9	45.52	14.33	71	0.23	025.41	0
025	Iron (ppm)	0505	472.4	1.600	459.9	45.52	14.33	71	0.28	025.31	0
025	Iron (ppm)	0878	474.0	4.000	459.9	45.52	14.33	71	0.31	025.41	0
025	Iron (ppm)	0083	474.0	2.000	459.9	45.52	14.33	71	0.31	025.43	0
025	Iron (ppm)	0098	475.4	12.30	459.9	45.52	14.33	71	0.34	025.41	0
025	Iron (ppm)	0939	476.5	11.00	459.9	45.52	14.33	71	0.37	025.53	0
025	Iron (ppm)	0100	476.5	7.000	459.9	45.52	14.33	71	0.37	025.99	0
025	Iron (ppm)	0353	477.0	4.000	459.9	45.52	14.33	71	0.38	025.43	0
025	Iron (ppm)	0510	477.0	2.000	459.9	45.52	14.33	71	0.38	025.43	0
025	Iron (ppm)	0918	479.1	11.79	459.9	45.52	14.33	71	0.42	025.43	0
025	Iron (ppm)	0407	484.2	5.894	459.9	45.52	14.33	71	0.53	025.43	0
025	Iron (ppm)	0229	500.3	1.370	459.9	45.52	14.33	71	0.89	025.41	0
025	Iron (ppm)	0035	500.5	11.00	459.9	45.52	14.33	71	0.89	025.42	0
025	Iron (ppm)	0939	503.8	12.45	459.9	45.52	14.33	71	0.96	025.31	0
025	Iron (ppm)	0208	504.2	9.000	459.9	45.52	14.33	71	0.97	025.41	0
025	Iron (ppm)	0208	504.5	5.000	459.9	45.52	14.33	71	0.98	025.31	0
025	Iron (ppm)	0098	510.3	32.10	459.9	45.52	14.33	71	1.11	025.43	0
025	Iron (ppm)	0045	510.5	1.000	459.9	45.52	14.33	71	1.11	025.42	0
025	Iron (ppm)	2146	513.3	1.500	459.9	45.52	14.33	71	1.17	025.51	0
025	Iron (ppm)	0682	514.0	0.0000	459.9	45.52	14.33	71	1.19	025.43	0
025	Iron (ppm)	0035	514.5	21.00	459.9	45.52	14.33	71	1.20	025.43	0
025	Iron (ppm)	0536	525.9	10.00	459.9	45.52	14.33	71	1.45	025.31	0
025	Iron (ppm)	0884	537.1	1.600	459.9	45.52	14.33	71	1.70	025.31	0
025	Iron (ppm)	2188	539.2	1.740	459.9	45.52	14.33	71	1.74	025.31	0
025	Iron (ppm)	0032	556.5	6.200	459.9	45.52	14.33	71	2.12	025.43	0
025	Iron (ppm)	2062	592.0	18.00	459.9	45.52	14.33	71	2.90	025.31	0
025	Iron (ppm)	2089	600.0	58.91	459.9	45.52	14.33	71	3.08	025.43	0
025	Iron (ppm)	0123	213.1	417.8	459.9	45.52	14.33	71	-5.42	025.41	1
027	Magnesium (%)	0294	0.2100	0.0000	0.2449	0.0131	0.0070	78	-2.66	027.42	0
027	Magnesium (%)	0089	0.2200	0.0000	0.2449	0.0131	0.0070	78	-1.90	027.43	0
027	Magnesium (%)	2129	0.2219	0.0031	0.2449	0.0131	0.0070	78	-1.76	027.42	0
027	Magnesium (%)	0560	0.2236	0.0287	0.2449	0.0131	0.0070	78	-1.63	027.52	0
027	Magnesium (%)	0407	0.2250	0.0019	0.2449	0.0131	0.0070	78	-1.52	027.41	0
027	Magnesium (%)	0019	0.2250	0.0300	0.2449	0.0131	0.0070	78	-1.52	027.41	0
027	Magnesium (%)	0096	0.2255	0.0110	0.2449	0.0131	0.0070	78	-1.48	027.42	0
027	Magnesium (%)	0229	0.2300	0.0200	0.2449	0.0131	0.0070	78	-1.14	027.41	0
027	Magnesium (%)	2053	0.2300	0.0200	0.2449	0.0131	0.0070	78	-1.14	027.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
027	Magnesium (%)	0175	0.2300	0.0000	0.2449	0.0131	0.0070	78	-1.14	027.31	0
027	Magnesium (%)	0910	0.2300	0.0000	0.2449	0.0131	0.0070	78	-1.14	027.41	0
027	Magnesium (%)	0035	0.2300	0.0000	0.2449	0.0131	0.0070	78	-1.14	027.43	0
027	Magnesium (%)	0083	0.2300	0.0000	0.2449	0.0131	0.0070	78	-1.14	027.43	0
027	Magnesium (%)	0202	0.2305	0.0050	0.2449	0.0131	0.0070	78	-1.10	027.43	0
027	Magnesium (%)	0948	0.2325	0.0050	0.2449	0.0131	0.0070	78	-0.95	027.33	0
027	Magnesium (%)	0941	0.2340	0.0000	0.2449	0.0131	0.0070	78	-0.83	027.43	0
027	Magnesium (%)	0357	0.2353	0.0032	0.2449	0.0131	0.0070	78	-0.73	027.42	0
027	Magnesium (%)	0186	0.2379	0.0083	0.2449	0.0131	0.0070	78	-0.54	027.42	0
027	Magnesium (%)	0407	0.2388	0.0145	0.2449	0.0131	0.0070	78	-0.47	027.43	0
027	Magnesium (%)	0870	0.2392	0.0037	0.2449	0.0131	0.0070	78	-0.44	027.42	0
027	Magnesium (%)	0148	0.2395	0.0170	0.2449	0.0131	0.0070	78	-0.41	027.41	0
027	Magnesium (%)	0345	0.2395	0.0050	0.2449	0.0131	0.0070	78	-0.41	027.43	0
027	Magnesium (%)	0520	0.2400	0.0000	0.2449	0.0131	0.0070	78	-0.37	027.41	0
027	Magnesium (%)	2012	0.2400	0.0000	0.2449	0.0131	0.0070	78	-0.37	027.41	0
027	Magnesium (%)	0035	0.2400	0.0000	0.2449	0.0131	0.0070	78	-0.37	027.42	0
027	Magnesium (%)	0265	0.2400	0.0000	0.2449	0.0131	0.0070	78	-0.37	027.42	0
027	Magnesium (%)	0278	0.2400	0.0200	0.2449	0.0131	0.0070	78	-0.37	027.42	0
027	Magnesium (%)	0726	0.2400	0.0000	0.2449	0.0131	0.0070	78	-0.37	027.42	0
027	Magnesium (%)	0353	0.2400	0.0200	0.2449	0.0131	0.0070	78	-0.37	027.43	0
027	Magnesium (%)	0510	0.2400	0.0000	0.2449	0.0131	0.0070	78	-0.37	027.43	0
027	Magnesium (%)	0199	0.2400	0.0000	0.2449	0.0131	0.0070	78	-0.37	027.53	0
027	Magnesium (%)	0154	0.2411	0.0079	0.2449	0.0131	0.0070	78	-0.29	027.52	0
027	Magnesium (%)	0263	0.2415	0.0010	0.2449	0.0131	0.0070	78	-0.26	027.42	0
027	Magnesium (%)	0171	0.2420	0.0020	0.2449	0.0131	0.0070	78	-0.22	027.41	0
027	Magnesium (%)	0142	0.2425	0.0017	0.2449	0.0131	0.0070	78	-0.19	027.31	0
027	Magnesium (%)	0968	0.2430	0.0020	0.2449	0.0131	0.0070	78	-0.14	027.43	0
027	Magnesium (%)	0186	0.2444	0.0032	0.2449	0.0131	0.0070	78	-0.04	027.52	0
027	Magnesium (%)	0139	0.2448	0.0090	0.2449	0.0131	0.0070	78	-0.01	027.41	0
027	Magnesium (%)	0868	0.2450	0.0100	0.2449	0.0131	0.0070	78	0.01	027.31	0
027	Magnesium (%)	0169	0.2450	0.0100	0.2449	0.0131	0.0070	78	0.01	027.32	0
027	Magnesium (%)	0226	0.2450	0.0100	0.2449	0.0131	0.0070	78	0.01	027.41	0
027	Magnesium (%)	0878	0.2450	0.0100	0.2449	0.0131	0.0070	78	0.01	027.41	0
027	Magnesium (%)	0045	0.2450	0.0020	0.2449	0.0131	0.0070	78	0.01	027.42	0
027	Magnesium (%)	0939	0.2450	0.0300	0.2449	0.0131	0.0070	78	0.01	027.53	0
027	Magnesium (%)	0590	0.2450	0.0100	0.2449	0.0131	0.0070	78	0.01	027.99	0
027	Magnesium (%)	0598	0.2455	0.0043	0.2449	0.0131	0.0070	78	0.04	027.41	0
027	Magnesium (%)	0074	0.2455	0.0010	0.2449	0.0131	0.0070	78	0.05	027.41	0
027	Magnesium (%)	0096	0.2475	0.0230	0.2449	0.0131	0.0070	78	0.20	027.52	0
027	Magnesium (%)	0572	0.2475	0.0310	0.2449	0.0131	0.0070	78	0.20	027.53	0
027	Magnesium (%)	0037	0.2495	0.0070	0.2449	0.0131	0.0070	78	0.35	027.42	0
027	Magnesium (%)	0536	0.2500	0.0000	0.2449	0.0131	0.0070	78	0.39	027.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
027	Magnesium (%)	0884	0.2500	0.0000	0.2449	0.0131	0.0070	78	0.39	027.31	0
027	Magnesium (%)	0003	0.2500	0.0000	0.2449	0.0131	0.0070	78	0.39	027.41	0
027	Magnesium (%)	0354	0.2500	0.0000	0.2449	0.0131	0.0070	78	0.39	027.41	0
027	Magnesium (%)	0692	0.2500	0.0200	0.2449	0.0131	0.0070	78	0.39	027.42	0
027	Magnesium (%)	0297	0.2500	0.0000	0.2449	0.0131	0.0070	78	0.39	027.43	0
027	Magnesium (%)	0675	0.2500	0.0000	0.2449	0.0131	0.0070	78	0.39	027.43	0
027	Magnesium (%)	0164	0.2520	0.0020	0.2449	0.0131	0.0070	78	0.54	027.44	0
027	Magnesium (%)	2146	0.2550	0.0300	0.2449	0.0131	0.0070	78	0.77	027.31	0
027	Magnesium (%)	0358	0.2550	0.0100	0.2449	0.0131	0.0070	78	0.77	027.41	0
027	Magnesium (%)	0190	0.2550	0.0100	0.2449	0.0131	0.0070	78	0.77	027.42	0
027	Magnesium (%)	0098	0.2550	0.0100	0.2449	0.0131	0.0070	78	0.77	027.43	0
027	Magnesium (%)	0208	0.2565	0.0090	0.2449	0.0131	0.0070	78	0.89	027.41	0
027	Magnesium (%)	0208	0.2585	0.0070	0.2449	0.0131	0.0070	78	1.04	027.31	0
027	Magnesium (%)	0553	0.2585	0.0050	0.2449	0.0131	0.0070	78	1.04	027.53	0
027	Magnesium (%)	0939	0.2600	0.0000	0.2449	0.0131	0.0070	78	1.15	027.31	0
027	Magnesium (%)	2196	0.2600	0.0000	0.2449	0.0131	0.0070	78	1.15	027.31	0
027	Magnesium (%)	0425	0.2600	0.0000	0.2449	0.0131	0.0070	78	1.15	027.43	0
027	Magnesium (%)	2089	0.2600	0.0000	0.2449	0.0131	0.0070	78	1.15	027.43	0
027	Magnesium (%)	0100	0.2600	0.0000	0.2449	0.0131	0.0070	78	1.15	027.99	0
027	Magnesium (%)	0032	0.2609	0.0004	0.2449	0.0131	0.0070	78	1.22	027.43	0
027	Magnesium (%)	0529	0.2640	0.0140	0.2449	0.0131	0.0070	78	1.46	027.31	0
027	Magnesium (%)	0918	0.2640	0.0000	0.2449	0.0131	0.0070	78	1.46	027.43	0
027	Magnesium (%)	0693	0.2690	0.0100	0.2449	0.0131	0.0070	78	1.84	027.42	0
027	Magnesium (%)	0098	0.2700	0.0200	0.2449	0.0131	0.0070	78	1.92	027.41	0
027	Magnesium (%)	2146	0.2735	0.0030	0.2449	0.0131	0.0070	78	2.19	027.51	0
027	Magnesium (%)	0682	0.2800	0.0000	0.2449	0.0131	0.0070	78	2.68	027.43	0
027	Magnesium (%)	0042	0.2865	0.0030	0.2449	0.0131	0.0070	78	3.18	027.42	0
027	Magnesium (%)	2302	0.4050	0.0100	0.2449	0.0131	0.0070	78	12.22	027.99	2
028	Manganese (ppm)	2234	116.7	1.650	212.7	13.77	6.446	84	-6.97	028.43	0
028	Manganese (ppm)	0199	132.2	0.0600	212.7	13.77	6.446	84	-5.84	028.53	0
028	Manganese (ppm)	2062	148.1	0.7900	212.7	13.77	6.446	84	-4.69	028.31	0
028	Manganese (ppm)	0226	163.7	9.150	212.7	13.77	6.446	84	-3.56	028.41	0
028	Manganese (ppm)	0354	172.0	6.040	212.7	13.77	6.446	84	-2.96	028.41	0
028	Manganese (ppm)	0066	172.7	4.280	212.7	13.77	6.446	84	-2.90	028.44	0
028	Manganese (ppm)	0009	189.5	7.000	212.7	13.77	6.446	84	-1.68	028.42	0
028	Manganese (ppm)	0520	193.5	1.000	212.7	13.77	6.446	84	-1.39	028.41	0
028	Manganese (ppm)	0939	193.5	11.00	212.7	13.77	6.446	84	-1.39	028.53	0
028	Manganese (ppm)	0096	195.7	17.29	212.7	13.77	6.446	84	-1.23	028.42	0
028	Manganese (ppm)	0948	196.8	10.01	212.7	13.77	6.446	84	-1.15	028.33	0
028	Manganese (ppm)	0918	198.0	5.000	212.7	13.77	6.446	84	-1.07	028.43	0
028	Manganese (ppm)	0074	198.0	6.000	212.7	13.77	6.446	84	-1.06	028.41	0
028	Manganese (ppm)	0278	199.5	5.000	212.7	13.77	6.446	84	-0.96	028.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
028	Manganese (ppm)	0553	199.5	11.00	212.7	13.77	6.446	84	-0.96	028.53	0
028	Manganese (ppm)	0263	199.8	0.5200	212.7	13.77	6.446	84	-0.93	028.42	0
028	Manganese (ppm)	2053	200.0	2.000	212.7	13.77	6.446	84	-0.92	028.42	0
028	Manganese (ppm)	0910	202.0	6.000	212.7	13.77	6.446	84	-0.77	028.41	0
028	Manganese (ppm)	2129	202.7	0.6000	212.7	13.77	6.446	84	-0.72	028.42	0
028	Manganese (ppm)	0505	202.9	20.70	212.7	13.77	6.446	84	-0.71	028.43	0
028	Manganese (ppm)	0870	202.9	11.00	212.7	13.77	6.446	84	-0.71	028.42	0
028	Manganese (ppm)	0510	203.0	2.000	212.7	13.77	6.446	84	-0.70	028.43	0
028	Manganese (ppm)	2302	204.0	20.00	212.7	13.77	6.446	84	-0.63	028.99	0
028	Manganese (ppm)	2012	204.5	2.290	212.7	13.77	6.446	84	-0.59	028.41	0
028	Manganese (ppm)	0968	205.0	2.000	212.7	13.77	6.446	84	-0.56	028.43	0
028	Manganese (ppm)	0693	205.4	9.146	212.7	13.77	6.446	84	-0.53	028.42	0
028	Manganese (ppm)	0939	207.0	4.000	212.7	13.77	6.446	84	-0.41	028.31	0
028	Manganese (ppm)	0353	207.0	16.00	212.7	13.77	6.446	84	-0.41	028.43	0
028	Manganese (ppm)	0921	208.0	7.600	212.7	13.77	6.446	84	-0.34	028.31	0
028	Manganese (ppm)	0536	208.5	1.820	212.7	13.77	6.446	84	-0.30	028.31	0
028	Manganese (ppm)	0878	209.5	3.000	212.7	13.77	6.446	84	-0.23	028.41	0
028	Manganese (ppm)	0169	209.5	1.000	212.7	13.77	6.446	84	-0.23	028.43	0
028	Manganese (ppm)	2146	209.9	14.47	212.7	13.77	6.446	84	-0.20	028.31	0
028	Manganese (ppm)	0358	211.4	19.32	212.7	13.77	6.446	84	-0.09	028.41	0
028	Manganese (ppm)	0003	211.5	17.00	212.7	13.77	6.446	84	-0.08	028.41	0
028	Manganese (ppm)	0171	211.5	1.000	212.7	13.77	6.446	84	-0.08	028.41	0
028	Manganese (ppm)	0186	211.5	1.000	212.7	13.77	6.446	84	-0.08	028.42	0
028	Manganese (ppm)	0529	211.7	1.300	212.7	13.77	6.446	84	-0.07	028.31	0
028	Manganese (ppm)	0560	211.7	14.60	212.7	13.77	6.446	84	-0.07	028.52	0
028	Manganese (ppm)	0407	212.0	6.730	212.7	13.77	6.446	84	-0.05	028.41	0
028	Manganese (ppm)	0164	212.0	8.000	212.7	13.77	6.446	84	-0.05	028.44	0
028	Manganese (ppm)	0590	212.0	4.000	212.7	13.77	6.446	84	-0.05	028.99	0
028	Manganese (ppm)	0345	212.3	2.500	212.7	13.77	6.446	84	-0.03	028.43	0
028	Manganese (ppm)	0001	212.3	2.600	212.7	13.77	6.446	84	-0.03	028.31	0
028	Manganese (ppm)	0229	212.4	4.510	212.7	13.77	6.446	84	-0.02	028.41	0
028	Manganese (ppm)	2196	213.7	0.0000	212.7	13.77	6.446	84	0.07	028.31	0
028	Manganese (ppm)	0035	214.0	8.000	212.7	13.77	6.446	84	0.10	028.42	0
028	Manganese (ppm)	0675	214.7	6.720	212.7	13.77	6.446	84	0.15	028.43	0
028	Manganese (ppm)	0083	215.0	2.000	212.7	13.77	6.446	84	0.17	028.43	0
028	Manganese (ppm)	0019	215.3	17.10	212.7	13.77	6.446	84	0.19	028.41	0
028	Manganese (ppm)	0692	215.5	23.00	212.7	13.77	6.446	84	0.21	028.42	0
028	Manganese (ppm)	0297	216.0	12.00	212.7	13.77	6.446	84	0.24	028.43	0
028	Manganese (ppm)	0175	217.0	2.000	212.7	13.77	6.446	84	0.32	028.31	0
028	Manganese (ppm)	0208	217.5	3.000	212.7	13.77	6.446	84	0.35	028.31	0
028	Manganese (ppm)	0148	218.0	14.00	212.7	13.77	6.446	84	0.39	028.41	0
028	Manganese (ppm)	0265	218.0	6.000	212.7	13.77	6.446	84	0.39	028.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
028	Manganese (ppm)	0208	218.4	6.900	212.7	13.77	6.446	84	0.41	028.41	0
028	Manganese (ppm)	0035	219.0	20.00	212.7	13.77	6.446	84	0.46	028.43	0
028	Manganese (ppm)	0884	219.6	7.400	212.7	13.77	6.446	84	0.50	028.31	0
028	Manganese (ppm)	0572	221.0	14.00	212.7	13.77	6.446	84	0.61	028.53	0
028	Manganese (ppm)	0598	221.3	2.600	212.7	13.77	6.446	84	0.63	028.41	0
028	Manganese (ppm)	0098	221.4	10.90	212.7	13.77	6.446	84	0.63	028.41	0
028	Manganese (ppm)	0045	221.5	1.000	212.7	13.77	6.446	84	0.64	028.42	0
028	Manganese (ppm)	0096	222.3	8.613	212.7	13.77	6.446	84	0.70	028.52	0
028	Manganese (ppm)	0868	222.5	11.00	212.7	13.77	6.446	84	0.72	028.31	0
028	Manganese (ppm)	0656	224.0	0.0000	212.7	13.77	6.446	84	0.82	028.31	0
028	Manganese (ppm)	0186	224.0	10.00	212.7	13.77	6.446	84	0.82	028.52	0
028	Manganese (ppm)	0682	224.7	0.0000	212.7	13.77	6.446	84	0.88	028.43	0
028	Manganese (ppm)	0941	225.0	0.0000	212.7	13.77	6.446	84	0.90	028.43	0
028	Manganese (ppm)	0100	225.0	2.000	212.7	13.77	6.446	84	0.90	028.99	0
028	Manganese (ppm)	0098	225.5	2.800	212.7	13.77	6.446	84	0.93	028.43	0
028	Manganese (ppm)	0202	226.7	11.98	212.7	13.77	6.446	84	1.02	028.43	0
028	Manganese (ppm)	2146	227.7	4.500	212.7	13.77	6.446	84	1.09	028.51	0
028	Manganese (ppm)	0407	228.8	1.350	212.7	13.77	6.446	84	1.17	028.43	0
028	Manganese (ppm)	0425	228.9	3.000	212.7	13.77	6.446	84	1.18	028.43	0
028	Manganese (ppm)	0190	229.4	1.000	212.7	13.77	6.446	84	1.22	028.42	0
028	Manganese (ppm)	0357	231.0	2.436	212.7	13.77	6.446	84	1.33	028.42	0
028	Manganese (ppm)	2089	231.6	0.3200	212.7	13.77	6.446	84	1.37	028.43	0
028	Manganese (ppm)	0726	231.7	0.2000	212.7	13.77	6.446	84	1.38	028.42	0
028	Manganese (ppm)	0032	232.1	0.6000	212.7	13.77	6.446	84	1.41	028.43	0
028	Manganese (ppm)	0037	235.2	0.5000	212.7	13.77	6.446	84	1.63	028.42	0
028	Manganese (ppm)	0042	235.5	21.00	212.7	13.77	6.446	84	1.66	028.42	0
028	Manganese (ppm)	0294	240.3	1.120	212.7	13.77	6.446	84	2.01	028.42	0
028	Manganese (ppm)	2188	262.3	0.4700	212.7	13.77	6.446	84	3.61	028.31	0
031	Phosphorus (%)	0921	0.5535	0.1150	0.6757	0.0347	0.0185	111	-3.52	031.01	0
031	Phosphorus (%)	0089	0.5550	0.0100	0.6757	0.0347	0.0185	111	-3.48	031.43	0
031	Phosphorus (%)	0009	0.5640	0.0099	0.6757	0.0347	0.0185	111	-3.22	031.42	0
031	Phosphorus (%)	0560	0.5769	0.0384	0.6757	0.0347	0.0185	111	-2.85	031.52	0
031	Phosphorus (%)	0848	0.5850	0.0100	0.6757	0.0347	0.0185	111	-2.61	031.41	0
031	Phosphorus (%)	0199	0.5950	0.0300	0.6757	0.0347	0.0185	111	-2.32	031.53	0
031	Phosphorus (%)	0154	0.5972	0.0776	0.6757	0.0347	0.0185	111	-2.26	031.52	0
031	Phosphorus (%)	0278	0.6000	0.0400	0.6757	0.0347	0.0185	111	-2.18	031.42	0
031	Phosphorus (%)	0202	0.6020	0.0000	0.6757	0.0347	0.0185	111	-2.12	031.43	0
031	Phosphorus (%)	0019	0.6150	0.0500	0.6757	0.0347	0.0185	111	-1.75	031.41	0
031	Phosphorus (%)	2302	0.6250	0.1900	0.6757	0.0347	0.0185	111	-1.46	031.99	0
031	Phosphorus (%)	0096	0.6260	0.0280	0.6757	0.0347	0.0185	111	-1.43	031.42	0
031	Phosphorus (%)	2259	0.6265	0.0030	0.6757	0.0347	0.0185	111	-1.42	031.41	0
031	Phosphorus (%)	0186	0.6272	0.0118	0.6757	0.0347	0.0185	111	-1.40	031.42	0



Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	0910	0.6300	0.0200	0.6757	0.0347	0.0185	111	-1.32	031.41	0
031	Phosphorus (%)	0265	0.6400	0.0000	0.6757	0.0347	0.0185	111	-1.03	031.42	0
031	Phosphorus (%)	0692	0.6400	0.0200	0.6757	0.0347	0.0185	111	-1.03	031.42	0
031	Phosphorus (%)	0676	0.6400	0.0000	0.6757	0.0347	0.0185	111	-1.03	031.99	0
031	Phosphorus (%)	0186	0.6427	0.0280	0.6757	0.0347	0.0185	111	-0.95	031.52	0
031	Phosphorus (%)	0941	0.6430	0.0000	0.6757	0.0347	0.0185	111	-0.94	031.43	0
031	Phosphorus (%)	0175	0.6450	0.0100	0.6757	0.0347	0.0185	111	-0.88	031.01	0
031	Phosphorus (%)	2295	0.6450	0.0100	0.6757	0.0347	0.0185	111	-0.88	031.01	0
031	Phosphorus (%)	0294	0.6450	0.0100	0.6757	0.0347	0.0185	111	-0.88	031.42	0
031	Phosphorus (%)	0870	0.6490	0.0221	0.6757	0.0347	0.0185	111	-0.77	031.42	0
031	Phosphorus (%)	0687	0.6500	0.1000	0.6757	0.0347	0.0185	111	-0.74	031.01	0
031	Phosphorus (%)	2146	0.6500	0.0200	0.6757	0.0347	0.0185	111	-0.74	031.01	0
031	Phosphorus (%)	2188	0.6500	0.0000	0.6757	0.0347	0.0185	111	-0.74	031.01	0
031	Phosphorus (%)	2129	0.6504	0.0122	0.6757	0.0347	0.0185	111	-0.73	031.42	0
031	Phosphorus (%)	0939	0.6550	0.0100	0.6757	0.0347	0.0185	111	-0.60	031.01	0
031	Phosphorus (%)	0520	0.6550	0.0100	0.6757	0.0347	0.0185	111	-0.60	031.41	0
031	Phosphorus (%)	0100	0.6550	0.0100	0.6757	0.0347	0.0185	111	-0.60	031.99	0
031	Phosphorus (%)	0572	0.6555	0.0010	0.6757	0.0347	0.0185	111	-0.58	031.53	0
031	Phosphorus (%)	0693	0.6585	0.0130	0.6757	0.0347	0.0185	111	-0.50	031.42	0
031	Phosphorus (%)	2144	0.6600	0.0000	0.6757	0.0347	0.0185	111	-0.45	031.01	0
031	Phosphorus (%)	0208	0.6615	0.0270	0.6757	0.0347	0.0185	111	-0.41	031.01	0
031	Phosphorus (%)	0001	0.6645	0.0090	0.6757	0.0347	0.0185	111	-0.32	031.43	0
031	Phosphorus (%)	0035	0.6650	0.0100	0.6757	0.0347	0.0185	111	-0.31	031.42	0
031	Phosphorus (%)	0425	0.6650	0.0700	0.6757	0.0347	0.0185	111	-0.31	031.43	0
031	Phosphorus (%)	0939	0.6650	0.0300	0.6757	0.0347	0.0185	111	-0.31	031.53	0
031	Phosphorus (%)	0036	0.6683	0.0092	0.6757	0.0347	0.0185	111	-0.21	031.03	0
031	Phosphorus (%)	0139	0.6695	0.0010	0.6757	0.0347	0.0185	111	-0.18	031.41	0
031	Phosphorus (%)	2192	0.6700	0.0200	0.6757	0.0347	0.0185	111	-0.16	031.01	0
031	Phosphorus (%)	0229	0.6700	0.0000	0.6757	0.0347	0.0185	111	-0.16	031.41	0
031	Phosphorus (%)	0298	0.6700	0.0200	0.6757	0.0347	0.0185	111	-0.16	031.41	0
031	Phosphorus (%)	0010	0.6700	0.0000	0.6757	0.0347	0.0185	111	-0.16	031.43	0
031	Phosphorus (%)	0297	0.6700	0.0200	0.6757	0.0347	0.0185	111	-0.16	031.43	0
031	Phosphorus (%)	0164	0.6715	0.0030	0.6757	0.0347	0.0185	111	-0.12	031.44	0
031	Phosphorus (%)	0037	0.6725	0.0030	0.6757	0.0347	0.0185	111	-0.09	031.42	0
031	Phosphorus (%)	0263	0.6730	0.0040	0.6757	0.0347	0.0185	111	-0.08	031.42	0
031	Phosphorus (%)	0968	0.6730	0.0060	0.6757	0.0347	0.0185	111	-0.08	031.43	0
031	Phosphorus (%)	0553	0.6745	0.0070	0.6757	0.0347	0.0185	111	-0.03	031.53	0
031	Phosphorus (%)	0529	0.6750	0.0100	0.6757	0.0347	0.0185	111	-0.02	031.01	0
031	Phosphorus (%)	0868	0.6750	0.0100	0.6757	0.0347	0.0185	111	-0.02	031.01	0
031	Phosphorus (%)	0169	0.6750	0.0100	0.6757	0.0347	0.0185	111	-0.02	031.03	0
031	Phosphorus (%)	0083	0.6750	0.0100	0.6757	0.0347	0.0185	111	-0.02	031.43	0
031	Phosphorus (%)	0148	0.6765	0.0610	0.6757	0.0347	0.0185	111	0.02	031.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	2062	0.6782	0.0135	0.6757	0.0347	0.0185	111	0.07	031.01	0
031	Phosphorus (%)	2109	0.6790	0.0140	0.6757	0.0347	0.0185	111	0.10	031.01	0
031	Phosphorus (%)	0142	0.6792	0.0228	0.6757	0.0347	0.0185	111	0.10	031.01	0
031	Phosphorus (%)	2012	0.6800	0.0000	0.6757	0.0347	0.0185	111	0.12	031.41	0
031	Phosphorus (%)	0682	0.6800	0.0000	0.6757	0.0347	0.0185	111	0.12	031.43	0
031	Phosphorus (%)	0036	0.6804	0.0132	0.6757	0.0347	0.0185	111	0.14	031.43	0
031	Phosphorus (%)	0407	0.6822	0.0053	0.6757	0.0347	0.0185	111	0.19	031.41	0
031	Phosphorus (%)	0590	0.6850	0.0100	0.6757	0.0347	0.0185	111	0.27	031.01	0
031	Phosphorus (%)	2181	0.6850	0.0100	0.6757	0.0347	0.0185	111	0.27	031.01	0
031	Phosphorus (%)	0190	0.6850	0.0100	0.6757	0.0347	0.0185	111	0.27	031.42	0
031	Phosphorus (%)	2196	0.6870	0.0000	0.6757	0.0347	0.0185	111	0.33	031.01	0
031	Phosphorus (%)	0881	0.6880	0.0140	0.6757	0.0347	0.0185	111	0.35	031.01	0
031	Phosphorus (%)	0656	0.6900	0.0000	0.6757	0.0347	0.0185	111	0.41	031.01	0
031	Phosphorus (%)	0683	0.6900	0.0000	0.6757	0.0347	0.0185	111	0.41	031.01	0
031	Phosphorus (%)	0878	0.6900	0.0200	0.6757	0.0347	0.0185	111	0.41	031.01	0
031	Phosphorus (%)	0505	0.6900	0.0200	0.6757	0.0347	0.0185	111	0.41	031.02	0
031	Phosphorus (%)	0536	0.6900	0.0000	0.6757	0.0347	0.0185	111	0.41	031.06	0
031	Phosphorus (%)	0144	0.6900	0.0000	0.6757	0.0347	0.0185	111	0.41	031.41	0
031	Phosphorus (%)	0226	0.6900	0.0400	0.6757	0.0347	0.0185	111	0.41	031.41	0
031	Phosphorus (%)	0353	0.6900	0.0200	0.6757	0.0347	0.0185	111	0.41	031.43	0
031	Phosphorus (%)	2006	0.6910	0.0240	0.6757	0.0347	0.0185	111	0.44	031.01	0
031	Phosphorus (%)	0001	0.6912	0.0062	0.6757	0.0347	0.0185	111	0.45	031.03	0
031	Phosphorus (%)	0918	0.6940	0.0200	0.6757	0.0347	0.0185	111	0.53	031.43	0
031	Phosphorus (%)	0885	0.6941	0.0070	0.6757	0.0347	0.0185	111	0.53	031.01	0
031	Phosphorus (%)	0171	0.6950	0.0100	0.6757	0.0347	0.0185	111	0.56	031.41	0
031	Phosphorus (%)	0345	0.6960	0.0100	0.6757	0.0347	0.0185	111	0.58	031.43	0
031	Phosphorus (%)	0045	0.6980	0.0240	0.6757	0.0347	0.0185	111	0.64	031.42	0
031	Phosphorus (%)	0674	0.7000	0.0000	0.6757	0.0347	0.0185	111	0.70	031.01	0
031	Phosphorus (%)	0728	0.7000	0.0000	0.6757	0.0347	0.0185	111	0.70	031.01	0
031	Phosphorus (%)	0884	0.7000	0.0000	0.6757	0.0347	0.0185	111	0.70	031.01	0
031	Phosphorus (%)	0035	0.7000	0.0000	0.6757	0.0347	0.0185	111	0.70	031.43	0
031	Phosphorus (%)	0675	0.7000	0.0000	0.6757	0.0347	0.0185	111	0.70	031.43	0
031	Phosphorus (%)	0357	0.7007	0.0023	0.6757	0.0347	0.0185	111	0.72	031.42	0
031	Phosphorus (%)	0018	0.7050	0.0240	0.6757	0.0347	0.0185	111	0.84	031.01	0
031	Phosphorus (%)	0354	0.7050	0.0100	0.6757	0.0347	0.0185	111	0.84	031.41	0
031	Phosphorus (%)	0726	0.7050	0.0300	0.6757	0.0347	0.0185	111	0.84	031.42	0
031	Phosphorus (%)	2089	0.7050	0.0100	0.6757	0.0347	0.0185	111	0.84	031.43	0
031	Phosphorus (%)	2302	0.7050	0.0500	0.6757	0.0347	0.0185	111	0.84	031.01	0
031	Phosphorus (%)	0042	0.7070	0.0720	0.6757	0.0347	0.0185	111	0.90	031.42	0
031	Phosphorus (%)	0407	0.7074	0.0196	0.6757	0.0347	0.0185	111	0.91	031.43	0
031	Phosphorus (%)	2053	0.7100	0.0200	0.6757	0.0347	0.0185	111	0.99	031.02	0
031	Phosphorus (%)	0208	0.7100	0.0160	0.6757	0.0347	0.0185	111	0.99	031.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	0074	0.7125	0.0090	0.6757	0.0347	0.0185	111	1.06	031.41	0
031	Phosphorus (%)	0358	0.7200	0.0200	0.6757	0.0347	0.0185	111	1.28	031.41	0
031	Phosphorus (%)	0948	0.7220	0.0240	0.6757	0.0347	0.0185	111	1.33	031.01	0
031	Phosphorus (%)	0098	0.7230	0.0219	0.6757	0.0347	0.0185	111	1.36	031.43	0
031	Phosphorus (%)	0123	0.7250	0.0100	0.6757	0.0347	0.0185	111	1.42	031.41	0
031	Phosphorus (%)	0098	0.7275	0.0077	0.6757	0.0347	0.0185	111	1.49	031.41	0
031	Phosphorus (%)	0598	0.7285	0.0171	0.6757	0.0347	0.0185	111	1.52	031.41	0
031	Phosphorus (%)	0066	0.7310	0.0120	0.6757	0.0347	0.0185	111	1.59	031.44	0
031	Phosphorus (%)	0510	0.7350	0.0100	0.6757	0.0347	0.0185	111	1.71	031.43	0
031	Phosphorus (%)	0032	0.7489	0.0079	0.6757	0.0347	0.0185	111	2.11	031.43	0
031	Phosphorus (%)	0003	0.8900	0.0200	0.6757	0.0347	0.0185	111	6.17	031.41	0
031	Phosphorus (%)	0964	0.9138	0.0443	0.6757	0.0347	0.0185	111	6.86	031.43	0
031	Phosphorus (%)	2234	0.9450	0.0300	0.6757	0.0347	0.0185	111	7.76	031.43	0
031	Phosphorus (%)	0626	282.8	564.4	0.6757	0.0347	0.0185	111	8126.08	031.01	2
032	Potassium (%)	0089	0.8550	0.0100	1.092	0.0534	0.0261	76	-4.43	032.43	0
032	Potassium (%)	0154	0.9353	0.0121	1.092	0.0534	0.0261	76	-2.93	032.52	0
032	Potassium (%)	0529	0.9400	0.0800	1.092	0.0534	0.0261	76	-2.84	032.31	0
032	Potassium (%)	2302	0.9650	0.0300	1.092	0.0534	0.0261	76	-2.37	032.99	0
032	Potassium (%)	0019	0.9700	0.1000	1.092	0.0534	0.0261	76	-2.28	032.41	0
032	Potassium (%)	0683	0.9900	0.0000	1.092	0.0534	0.0261	76	-1.90	032.31	0
032	Potassium (%)	0148	0.9905	0.0190	1.092	0.0534	0.0261	76	-1.89	032.41	0
032	Potassium (%)	0560	1.018	0.1350	1.092	0.0534	0.0261	76	-1.39	032.52	0
032	Potassium (%)	2012	1.035	0.0100	1.092	0.0534	0.0261	76	-1.06	032.41	0
032	Potassium (%)	0083	1.035	0.0100	1.092	0.0534	0.0261	76	-1.06	032.43	0
032	Potassium (%)	2129	1.037	0.0230	1.092	0.0534	0.0261	76	-1.03	032.42	0
032	Potassium (%)	0870	1.039	0.0490	1.092	0.0534	0.0261	76	-0.99	032.42	0
032	Potassium (%)	0144	1.040	0.0200	1.092	0.0534	0.0261	76	-0.97	032.41	0
032	Potassium (%)	0354	1.040	0.0000	1.092	0.0534	0.0261	76	-0.97	032.41	0
032	Potassium (%)	0520	1.045	0.0100	1.092	0.0534	0.0261	76	-0.87	032.41	0
032	Potassium (%)	0353	1.045	0.0300	1.092	0.0534	0.0261	76	-0.87	032.43	0
032	Potassium (%)	0941	1.048	0.0000	1.092	0.0534	0.0261	76	-0.82	032.43	0
032	Potassium (%)	0675	1.050	0.0000	1.092	0.0534	0.0261	76	-0.78	032.43	0
032	Potassium (%)	0229	1.055	0.0500	1.092	0.0534	0.0261	76	-0.68	032.41	0
032	Potassium (%)	0199	1.055	0.0100	1.092	0.0534	0.0261	76	-0.68	032.53	0
032	Potassium (%)	0208	1.058	0.0700	1.092	0.0534	0.0261	76	-0.63	032.41	0
032	Potassium (%)	0939	1.060	0.0000	1.092	0.0534	0.0261	76	-0.59	032.31	0
032	Potassium (%)	0910	1.060	0.0000	1.092	0.0534	0.0261	76	-0.59	032.41	0
032	Potassium (%)	0868	1.065	0.0100	1.092	0.0534	0.0261	76	-0.50	032.31	0
032	Potassium (%)	0139	1.068	0.0010	1.092	0.0534	0.0261	76	-0.45	032.41	0
032	Potassium (%)	0357	1.068	0.0000	1.092	0.0534	0.0261	76	-0.44	032.42	0
032	Potassium (%)	0186	1.069	0.0230	1.092	0.0534	0.0261	76	-0.43	032.42	0
032	Potassium (%)	0278	1.070	0.0600	1.092	0.0534	0.0261	76	-0.40	032.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
032	Potassium (%)	0142	1.073	0.0342	1.092	0.0534	0.0261	76	-0.34	032.31	0
032	Potassium (%)	0345	1.079	0.0030	1.092	0.0534	0.0261	76	-0.24	032.43	0
032	Potassium (%)	0169	1.085	0.0300	1.092	0.0534	0.0261	76	-0.12	032.32	0
032	Potassium (%)	0918	1.085	0.0140	1.092	0.0534	0.0261	76	-0.12	032.43	0
032	Potassium (%)	0100	1.085	0.0100	1.092	0.0534	0.0261	76	-0.12	032.99	0
032	Potassium (%)	0186	1.087	0.0100	1.092	0.0534	0.0261	76	-0.09	032.52	0
032	Potassium (%)	0074	1.090	0.0000	1.092	0.0534	0.0261	76	-0.03	032.41	0
032	Potassium (%)	2089	1.090	0.0000	1.092	0.0534	0.0261	76	-0.03	032.43	0
032	Potassium (%)	0939	1.090	0.0200	1.092	0.0534	0.0261	76	-0.03	032.53	0
032	Potassium (%)	0263	1.091	0.0050	1.092	0.0534	0.0261	76	-0.02	032.42	0
032	Potassium (%)	0202	1.092	0.0110	1.092	0.0534	0.0261	76	0.00	032.43	0
032	Potassium (%)	0948	1.093	0.0130	1.092	0.0534	0.0261	76	0.02	032.31	0
032	Potassium (%)	0407	1.094	0.0112	1.092	0.0534	0.0261	76	0.05	032.41	0
032	Potassium (%)	0171	1.095	0.0100	1.092	0.0534	0.0261	76	0.06	032.41	0
032	Potassium (%)	0164	1.095	0.0300	1.092	0.0534	0.0261	76	0.06	032.44	0
032	Potassium (%)	0003	1.100	0.0000	1.092	0.0534	0.0261	76	0.16	032.41	0
032	Potassium (%)	0682	1.100	0.0000	1.092	0.0534	0.0261	76	0.16	032.43	0
032	Potassium (%)	0878	1.105	0.0100	1.092	0.0534	0.0261	76	0.25	032.41	0
032	Potassium (%)	0297	1.110	0.0600	1.092	0.0534	0.0261	76	0.35	032.43	0
032	Potassium (%)	0553	1.110	0.0000	1.092	0.0534	0.0261	76	0.35	032.53	0
032	Potassium (%)	0693	1.115	0.0750	1.092	0.0534	0.0261	76	0.43	032.42	0
032	Potassium (%)	0590	1.115	0.1100	1.092	0.0534	0.0261	76	0.44	032.99	0
032	Potassium (%)	0884	1.120	0.0000	1.092	0.0534	0.0261	76	0.53	032.02	0
032	Potassium (%)	0208	1.120	0.1000	1.092	0.0534	0.0261	76	0.53	032.31	0
032	Potassium (%)	0510	1.120	0.0000	1.092	0.0534	0.0261	76	0.53	032.43	0
032	Potassium (%)	0037	1.126	0.0060	1.092	0.0534	0.0261	76	0.64	032.42	0
032	Potassium (%)	0226	1.130	0.0000	1.092	0.0534	0.0261	76	0.72	032.41	0
032	Potassium (%)	0598	1.130	0.0060	1.092	0.0534	0.0261	76	0.72	032.41	0
032	Potassium (%)	0294	1.130	0.0000	1.092	0.0534	0.0261	76	0.72	032.42	0
032	Potassium (%)	0726	1.130	0.0000	1.092	0.0534	0.0261	76	0.72	032.42	0
032	Potassium (%)	0572	1.130	0.0400	1.092	0.0534	0.0261	76	0.72	032.53	0
032	Potassium (%)	0035	1.135	0.0300	1.092	0.0534	0.0261	76	0.81	032.43	0
032	Potassium (%)	2053	1.140	0.0800	1.092	0.0534	0.0261	76	0.91	032.31	0
032	Potassium (%)	0035	1.140	0.1000	1.092	0.0534	0.0261	76	0.91	032.42	0
032	Potassium (%)	0032	1.141	0.0050	1.092	0.0534	0.0261	76	0.92	032.43	0
032	Potassium (%)	0042	1.145	0.0700	1.092	0.0534	0.0261	76	1.00	032.42	0
032	Potassium (%)	0265	1.145	0.1100	1.092	0.0534	0.0261	76	1.00	032.42	0
032	Potassium (%)	0692	1.145	0.0500	1.092	0.0534	0.0261	76	1.00	032.42	0
032	Potassium (%)	0098	1.145	0.0100	1.092	0.0534	0.0261	76	1.00	032.43	0
032	Potassium (%)	0425	1.145	0.0100	1.092	0.0534	0.0261	76	1.00	032.43	0
032	Potassium (%)	0190	1.150	0.0200	1.092	0.0534	0.0261	76	1.09	032.42	0
032	Potassium (%)	0968	1.150	0.0020	1.092	0.0534	0.0261	76	1.09	032.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
032	Potassium (%)	0536	1.170	0.0200	1.092	0.0534	0.0261	76	1.47	032.31	0
032	Potassium (%)	0358	1.170	0.0200	1.092	0.0534	0.0261	76	1.47	032.41	0
032	Potassium (%)	0045	1.175	0.0300	1.092	0.0534	0.0261	76	1.56	032.42	0
032	Potassium (%)	0407	1.191	0.0141	1.092	0.0534	0.0261	76	1.86	032.43	0
032	Potassium (%)	0098	1.195	0.0300	1.092	0.0534	0.0261	76	1.94	032.41	0
032	Potassium (%)	0123	1.205	0.0100	1.092	0.0534	0.0261	76	2.12	032.41	0
032	Potassium (%)	0096	1.049	0.1430	1.092	0.0534	0.0261	76	-0.81	032.42	1
033	Salt as chloride (%)	0358	0.8400	0.0200	1.480	0.0846	0.0187	55	-7.57	033.99	0
033	Salt as chloride (%)	0171	0.8550	0.0100	1.480	0.0846	0.0187	55	-7.39	033.00	0
033	Salt as chloride (%)	0536	0.9750	0.0300	1.480	0.0846	0.0187	55	-5.97	033.99	0
033	Salt as chloride (%)	0425	1.050	0.0000	1.480	0.0846	0.0187	55	-5.09	033.01	0
033	Salt as chloride (%)	0353	1.240	0.0000	1.480	0.0846	0.0187	55	-2.84	033.00	0
033	Salt as chloride (%)	0505	1.260	0.0400	1.480	0.0846	0.0187	55	-2.61	033.03	0
033	Salt as chloride (%)	2076	1.310	0.1000	1.480	0.0846	0.0187	55	-2.01	033.00	0
033	Salt as chloride (%)	0144	1.315	0.1100	1.480	0.0846	0.0187	55	-1.96	033.03	0
033	Salt as chloride (%)	0265	1.330	0.0200	1.480	0.0846	0.0187	55	-1.78	033.03	0
033	Salt as chloride (%)	0868	1.370	0.0200	1.480	0.0846	0.0187	55	-1.31	033.05	0
033	Salt as chloride (%)	0190	1.380	0.0000	1.480	0.0846	0.0187	55	-1.19	033.03	0
033	Salt as chloride (%)	2295	1.405	0.0100	1.480	0.0846	0.0187	55	-0.89	033.00	0
033	Salt as chloride (%)	2073	1.420	0.0200	1.480	0.0846	0.0187	55	-0.71	033.99	0
033	Salt as chloride (%)	0675	1.425	0.0700	1.480	0.0846	0.0187	55	-0.66	033.00	0
033	Salt as chloride (%)	0298	1.430	0.0200	1.480	0.0846	0.0187	55	-0.60	033.00	0
033	Salt as chloride (%)	2181	1.430	0.0000	1.480	0.0846	0.0187	55	-0.60	033.99	0
033	Salt as chloride (%)	0910	1.435	0.0300	1.480	0.0846	0.0187	55	-0.54	033.99	0
033	Salt as chloride (%)	0885	1.442	0.0037	1.480	0.0846	0.0187	55	-0.45	033.00	0
033	Salt as chloride (%)	0045	1.460	0.0000	1.480	0.0846	0.0187	55	-0.24	033.00	0
033	Salt as chloride (%)	0884	1.465	0.0300	1.480	0.0846	0.0187	55	-0.18	033.00	0
033	Salt as chloride (%)	0208	1.490	0.0200	1.480	0.0846	0.0187	55	0.11	033.99	0
033	Salt as chloride (%)	0510	1.501	0.0330	1.480	0.0846	0.0187	55	0.24	033.01	0
033	Salt as chloride (%)	0083	1.505	0.0100	1.480	0.0846	0.0187	55	0.29	033.01	0
033	Salt as chloride (%)	0590	1.505	0.0700	1.480	0.0846	0.0187	55	0.29	033.01	0
033	Salt as chloride (%)	0683	1.505	0.0100	1.480	0.0846	0.0187	55	0.29	033.01	0
033	Salt as chloride (%)	0939	1.505	0.0100	1.480	0.0846	0.0187	55	0.29	033.01	0
033	Salt as chloride (%)	0878	1.510	0.0200	1.480	0.0846	0.0187	55	0.35	033.00	0
033	Salt as chloride (%)	0881	1.514	0.0130	1.480	0.0846	0.0187	55	0.39	033.00	0
033	Salt as chloride (%)	0309	1.514	0.0051	1.480	0.0846	0.0187	55	0.40	033.00	0
033	Salt as chloride (%)	0169	1.515	0.0300	1.480	0.0846	0.0187	55	0.41	033.00	0
033	Salt as chloride (%)	0914	1.515	0.0100	1.480	0.0846	0.0187	55	0.41	033.05	0
033	Salt as chloride (%)	0001	1.520	0.0000	1.480	0.0846	0.0187	55	0.47	033.01	0
033	Salt as chloride (%)	0098	1.520	0.0000	1.480	0.0846	0.0187	55	0.47	033.01	0
033	Salt as chloride (%)	0226	1.520	0.0000	1.480	0.0846	0.0187	55	0.47	033.01	0
033	Salt as chloride (%)	2146	1.520	0.0000	1.480	0.0846	0.0187	55	0.47	033.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
033	Salt as chloride (%)	0682	1.520	0.0000	1.480	0.0846	0.0187	55	0.47	033.99	0
033	Salt as chloride (%)	0175	1.525	0.0100	1.480	0.0846	0.0187	55	0.53	033.01	0
033	Salt as chloride (%)	0948	1.525	0.0100	1.480	0.0846	0.0187	55	0.53	033.01	0
033	Salt as chloride (%)	0407	1.527	0.0020	1.480	0.0846	0.0187	55	0.55	033.01	0
033	Salt as chloride (%)	0921	1.530	0.0200	1.480	0.0846	0.0187	55	0.59	033.00	0
033	Salt as chloride (%)	0148	1.530	0.0200	1.480	0.0846	0.0187	55	0.59	033.01	0
033	Salt as chloride (%)	0354	1.530	0.0000	1.480	0.0846	0.0187	55	0.59	033.01	0
033	Salt as chloride (%)	2109	1.531	0.0100	1.480	0.0846	0.0187	55	0.60	033.01	0
033	Salt as chloride (%)	2062	1.533	0.0252	1.480	0.0846	0.0187	55	0.62	033.00	0
033	Salt as chloride (%)	2146	1.535	0.0100	1.480	0.0846	0.0187	55	0.64	033.00	0
033	Salt as chloride (%)	2196	1.540	0.0000	1.480	0.0846	0.0187	55	0.70	033.00	0
033	Salt as chloride (%)	0164	1.550	0.0000	1.480	0.0846	0.0187	55	0.82	033.01	0
033	Salt as chloride (%)	0229	1.550	0.0400	1.480	0.0846	0.0187	55	0.82	033.01	0
033	Salt as chloride (%)	0726	1.550	0.0000	1.480	0.0846	0.0187	55	0.82	033.03	0
033	Salt as chloride (%)	0199	1.555	0.0100	1.480	0.0846	0.0187	55	0.88	033.01	0
033	Salt as chloride (%)	0693	1.557	0.0110	1.480	0.0846	0.0187	55	0.90	033.00	0
033	Salt as chloride (%)	0019	1.560	0.0000	1.480	0.0846	0.0187	55	0.94	033.01	0
033	Salt as chloride (%)	0100	1.570	0.0000	1.480	0.0846	0.0187	55	1.06	033.01	0
033	Salt as chloride (%)	0186	1.601	0.0470	1.480	0.0846	0.0187	55	1.42	033.01	0
033	Salt as chloride (%)	0278	1.605	0.0500	1.480	0.0846	0.0187	55	1.47	033.01	0
033	Salt as chloride (%)	2129	1.418	0.1390	1.480	0.0846	0.0187	55	-0.73	033.99	1
033	Salt as chloride (%)	0003	1.500	0.2000	1.480	0.0846	0.0187	55	0.23	033.01	1
034	Selenium (ppm)	2302	0.5500	0.9000	2.759	0.5562	0.1894	25	-3.97	034.99	0
034	Selenium (ppm)	0171	2.164	0.0140	2.759	0.5562	0.1894	25	-1.07	034.04	0
034	Selenium (ppm)	0148	2.165	0.3100	2.759	0.5562	0.1894	25	-1.07	034.41	0
034	Selenium (ppm)	0199	2.185	0.0700	2.759	0.5562	0.1894	25	-1.03	034.53	0
034	Selenium (ppm)	0164	2.290	0.1200	2.759	0.5562	0.1894	25	-0.84	034.53	0
034	Selenium (ppm)	0045	2.310	0.7200	2.759	0.5562	0.1894	25	-0.81	034.04	0
034	Selenium (ppm)	0910	2.380	0.0800	2.759	0.5562	0.1894	25	-0.68	034.52	0
034	Selenium (ppm)	0941	2.450	0.0000	2.759	0.5562	0.1894	25	-0.55	034.43	0
034	Selenium (ppm)	0939	2.530	0.1000	2.759	0.5562	0.1894	25	-0.41	034.53	0
034	Selenium (ppm)	0038	2.555	0.0900	2.759	0.5562	0.1894	25	-0.37	034.01	0
034	Selenium (ppm)	0096	2.574	0.0340	2.759	0.5562	0.1894	25	-0.33	034.52	0
034	Selenium (ppm)	0186	2.576	0.0270	2.759	0.5562	0.1894	25	-0.33	034.52	0
034	Selenium (ppm)	0553	2.675	0.2500	2.759	0.5562	0.1894	25	-0.15	034.53	0
034	Selenium (ppm)	0032	2.810	0.2800	2.759	0.5562	0.1894	25	0.09	034.53	0
034	Selenium (ppm)	0560	2.868	0.6130	2.759	0.5562	0.1894	25	0.20	034.53	0
034	Selenium (ppm)	0918	2.869	0.0170	2.759	0.5562	0.1894	25	0.20	034.53	0
034	Selenium (ppm)	0169	2.880	0.1000	2.759	0.5562	0.1894	25	0.22	034.04	0
034	Selenium (ppm)	0208	2.990	0.1800	2.759	0.5562	0.1894	25	0.42	034.52	0
034	Selenium (ppm)	2012	3.084	0.0250	2.759	0.5562	0.1894	25	0.58	034.53	0
034	Selenium (ppm)	0036	3.100	0.2000	2.759	0.5562	0.1894	25	0.61	034.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
034	Selenium (ppm)	0968	3.434	0.0350	2.759	0.5562	0.1894	25	1.21	034.43	0
034	Selenium (ppm)	0190	3.465	0.0900	2.759	0.5562	0.1894	25	1.27	034.99	0
034	Selenium (ppm)	0964	3.502	0.3502	2.759	0.5562	0.1894	25	1.34	034.43	0
034	Selenium (ppm)	0682	4.320	0.0000	2.759	0.5562	0.1894	25	2.81	034.43	0
034	Selenium (ppm)	0003	5.753	0.1300	2.759	0.5562	0.1894	25	5.38	034.53	0
034	Selenium (ppm)	0692	4.000	5.200	2.759	0.5562	0.1894	25	2.23	034.42	1
034	Selenium (ppm)	0096	< 20		2.759	0.5562	0.1894	25		034.42	5
035	Sodium (%)	2234	0.1600	0.0000	0.2912	0.0170	0.0087	83	-7.70	035.43	0
035	Sodium (%)	0536	0.2350	0.0100	0.2912	0.0170	0.0087	83	-3.30	035.31	0
035	Sodium (%)	0294	0.2500	0.0000	0.2912	0.0170	0.0087	83	-2.42	035.42	0
035	Sodium (%)	0175	0.2550	0.0300	0.2912	0.0170	0.0087	83	-2.13	035.31	0
035	Sodium (%)	0353	0.2574	0.0065	0.2912	0.0170	0.0087	83	-1.99	035.43	0
035	Sodium (%)	0096	0.2695	0.0170	0.2912	0.0170	0.0087	83	-1.27	035.42	0
035	Sodium (%)	0529	0.2700	0.0020	0.2912	0.0170	0.0087	83	-1.25	035.31	0
035	Sodium (%)	0019	0.2700	0.0200	0.2912	0.0170	0.0087	83	-1.25	035.41	0
035	Sodium (%)	0298	0.2700	0.0200	0.2912	0.0170	0.0087	83	-1.25	035.41	0
035	Sodium (%)	0939	0.2700	0.0000	0.2912	0.0170	0.0087	83	-1.25	035.53	0
035	Sodium (%)	0560	0.2722	0.0350	0.2912	0.0170	0.0087	83	-1.12	035.52	0
035	Sodium (%)	0683	0.2750	0.0100	0.2912	0.0170	0.0087	83	-0.95	035.31	0
035	Sodium (%)	0505	0.2750	0.0100	0.2912	0.0170	0.0087	83	-0.95	035.43	0
035	Sodium (%)	0199	0.2750	0.0100	0.2912	0.0170	0.0087	83	-0.95	035.53	0
035	Sodium (%)	0520	0.2763	0.0033	0.2912	0.0170	0.0087	83	-0.88	035.41	0
035	Sodium (%)	0154	0.2765	0.0116	0.2912	0.0170	0.0087	83	-0.86	035.52	0
035	Sodium (%)	0948	0.2775	0.0210	0.2912	0.0170	0.0087	83	-0.81	035.31	0
035	Sodium (%)	2259	0.2780	0.0060	0.2912	0.0170	0.0087	83	-0.78	035.41	0
035	Sodium (%)	0884	0.2800	0.0000	0.2912	0.0170	0.0087	83	-0.66	035.02	0
035	Sodium (%)	0144	0.2800	0.0000	0.2912	0.0170	0.0087	83	-0.66	035.41	0
035	Sodium (%)	0354	0.2800	0.0000	0.2912	0.0170	0.0087	83	-0.66	035.41	0
035	Sodium (%)	0910	0.2800	0.0200	0.2912	0.0170	0.0087	83	-0.66	035.41	0
035	Sodium (%)	2012	0.2800	0.0000	0.2912	0.0170	0.0087	83	-0.66	035.41	0
035	Sodium (%)	0035	0.2800	0.0000	0.2912	0.0170	0.0087	83	-0.66	035.42	0
035	Sodium (%)	0278	0.2800	0.0200	0.2912	0.0170	0.0087	83	-0.66	035.42	0
035	Sodium (%)	0035	0.2800	0.0000	0.2912	0.0170	0.0087	83	-0.66	035.43	0
035	Sodium (%)	2129	0.2801	0.0032	0.2912	0.0170	0.0087	83	-0.65	035.42	0
035	Sodium (%)	0345	0.2825	0.0030	0.2912	0.0170	0.0087	83	-0.51	035.43	0
035	Sodium (%)	0510	0.2830	0.0000	0.2912	0.0170	0.0087	83	-0.48	035.43	0
035	Sodium (%)	0941	0.2830	0.0000	0.2912	0.0170	0.0087	83	-0.48	035.43	0
035	Sodium (%)	0921	0.2830	0.0100	0.2912	0.0170	0.0087	83	-0.48	035.31	0
035	Sodium (%)	0939	0.2850	0.0100	0.2912	0.0170	0.0087	83	-0.37	035.31	0
035	Sodium (%)	2053	0.2850	0.0100	0.2912	0.0170	0.0087	83	-0.37	035.42	0
035	Sodium (%)	0868	0.2865	0.0190	0.2912	0.0170	0.0087	83	-0.28	035.01	0
035	Sodium (%)	0870	0.2865	0.0100	0.2912	0.0170	0.0087	83	-0.28	035.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
035	Sodium (%)	0164	0.2875	0.0030	0.2912	0.0170	0.0087	83	-0.22	035.41	0
035	Sodium (%)	0918	0.2895	0.0010	0.2912	0.0170	0.0087	83	-0.10	035.43	0
035	Sodium (%)	2196	0.2900	0.0000	0.2912	0.0170	0.0087	83	-0.07	035.31	0
035	Sodium (%)	0171	0.2900	0.0000	0.2912	0.0170	0.0087	83	-0.07	035.41	0
035	Sodium (%)	0083	0.2900	0.0000	0.2912	0.0170	0.0087	83	-0.07	035.43	0
035	Sodium (%)	0229	0.2900	0.0200	0.2912	0.0170	0.0087	83	-0.07	035.41	0
035	Sodium (%)	0160	0.2900	0.0400	0.2912	0.0170	0.0087	83	-0.07	035.99	0
035	Sodium (%)	0553	0.2925	0.0050	0.2912	0.0170	0.0087	83	0.07	035.53	0
035	Sodium (%)	0407	0.2941	0.0067	0.2912	0.0170	0.0087	83	0.17	035.41	0
035	Sodium (%)	0142	0.2947	0.0060	0.2912	0.0170	0.0087	83	0.20	035.31	0
035	Sodium (%)	0878	0.2950	0.0100	0.2912	0.0170	0.0087	83	0.22	035.41	0
035	Sodium (%)	0100	0.2950	0.0100	0.2912	0.0170	0.0087	83	0.22	035.99	0
035	Sodium (%)	0868	0.2950	0.0300	0.2912	0.0170	0.0087	83	0.22	035.31	0
035	Sodium (%)	0265	0.2950	0.0300	0.2912	0.0170	0.0087	83	0.22	035.42	0
035	Sodium (%)	0098	0.2950	0.0300	0.2912	0.0170	0.0087	83	0.22	035.43	0
035	Sodium (%)	0208	0.2955	0.0050	0.2912	0.0170	0.0087	83	0.25	035.31	0
035	Sodium (%)	0598	0.2959	0.0008	0.2912	0.0170	0.0087	83	0.27	035.41	0
035	Sodium (%)	0148	0.2960	0.0040	0.2912	0.0170	0.0087	83	0.28	035.41	0
035	Sodium (%)	0693	0.2975	0.0110	0.2912	0.0170	0.0087	83	0.37	035.42	0
035	Sodium (%)	0656	0.3000	0.0000	0.2912	0.0170	0.0087	83	0.51	035.31	0
035	Sodium (%)	0226	0.3000	0.0000	0.2912	0.0170	0.0087	83	0.51	035.41	0
035	Sodium (%)	0037	0.3000	0.0020	0.2912	0.0170	0.0087	83	0.51	035.42	0
035	Sodium (%)	0726	0.3000	0.0000	0.2912	0.0170	0.0087	83	0.51	035.42	0
035	Sodium (%)	0089	0.3000	0.0000	0.2912	0.0170	0.0087	83	0.51	035.43	0
035	Sodium (%)	0202	0.3000	0.0000	0.2912	0.0170	0.0087	83	0.51	035.43	0
035	Sodium (%)	0682	0.3000	0.0000	0.2912	0.0170	0.0087	83	0.51	035.43	0
035	Sodium (%)	2089	0.3000	0.0000	0.2912	0.0170	0.0087	83	0.51	035.43	0
035	Sodium (%)	0096	0.3020	0.0240	0.2912	0.0170	0.0087	83	0.63	035.52	0
035	Sodium (%)	0572	0.3035	0.0090	0.2912	0.0170	0.0087	83	0.72	035.53	0
035	Sodium (%)	0042	0.3040	0.0140	0.2912	0.0170	0.0087	83	0.75	035.42	0
035	Sodium (%)	0098	0.3050	0.0100	0.2912	0.0170	0.0087	83	0.81	035.41	0
035	Sodium (%)	0263	0.3050	0.0040	0.2912	0.0170	0.0087	83	0.81	035.42	0
035	Sodium (%)	0297	0.3050	0.0100	0.2912	0.0170	0.0087	83	0.81	035.43	0
035	Sodium (%)	0675	0.3050	0.0100	0.2912	0.0170	0.0087	83	0.81	035.43	0
035	Sodium (%)	0881	0.3050	0.0040	0.2912	0.0170	0.0087	83	0.81	035.43	0
035	Sodium (%)	0208	0.3085	0.0170	0.2912	0.0170	0.0087	83	1.01	035.41	0
035	Sodium (%)	0407	0.3094	0.0007	0.2912	0.0170	0.0087	83	1.06	035.43	0
035	Sodium (%)	0045	0.3130	0.0000	0.2912	0.0170	0.0087	83	1.28	035.42	0
035	Sodium (%)	0186	0.3149	0.0074	0.2912	0.0170	0.0087	83	1.39	035.52	0
035	Sodium (%)	0169	0.3150	0.0100	0.2912	0.0170	0.0087	83	1.39	035.32	0
035	Sodium (%)	0190	0.3150	0.0100	0.2912	0.0170	0.0087	83	1.39	035.42	0
035	Sodium (%)	0590	0.3150	0.0100	0.2912	0.0170	0.0087	83	1.39	035.99	0



Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
035	Sodium (%)	0186	0.3204	0.0026	0.2912	0.0170	0.0087	83	1.71	035.42	0
035	Sodium (%)	0425	0.3250	0.0100	0.2912	0.0170	0.0087	83	1.98	035.43	0
035	Sodium (%)	0358	0.3300	0.0000	0.2912	0.0170	0.0087	83	2.27	035.41	0
035	Sodium (%)	0692	0.3350	0.0300	0.2912	0.0170	0.0087	83	2.57	035.42	0
035	Sodium (%)	0139	0.3395	0.0010	0.2912	0.0170	0.0087	83	2.83	035.41	0
035	Sodium (%)	0066	0.3790	0.0100	0.2912	0.0170	0.0087	83	5.15	035.41	0
035	Sodium (%)	2302	0.7500	0.0600	0.2912	0.0170	0.0087	83	26.92	035.99	2
036	Sulfur (%)	2302	0.2750	0.0100	0.3371	0.0191	0.0093	42	-3.24	036.99	0
036	Sulfur (%)	0870	0.2980	0.0036	0.3371	0.0191	0.0093	42	-2.04	036.42	0
036	Sulfur (%)	0294	0.3100	0.0000	0.3371	0.0191	0.0093	42	-1.42	036.42	0
036	Sulfur (%)	0692	0.3100	0.0000	0.3371	0.0191	0.0093	42	-1.42	036.42	0
036	Sulfur (%)	0226	0.3150	0.0100	0.3371	0.0191	0.0093	42	-1.15	036.04	0
036	Sulfur (%)	2053	0.3150	0.0100	0.3371	0.0191	0.0093	42	-1.15	036.04	0
036	Sulfur (%)	2129	0.3190	0.0017	0.3371	0.0191	0.0093	42	-0.95	036.42	0
036	Sulfur (%)	0357	0.3211	0.0020	0.3371	0.0191	0.0093	42	-0.84	036.42	0
036	Sulfur (%)	0941	0.3230	0.0000	0.3371	0.0191	0.0093	42	-0.74	036.43	0
036	Sulfur (%)	0186	0.3231	0.0067	0.3371	0.0191	0.0093	42	-0.73	036.42	0
036	Sulfur (%)	0693	0.3240	0.0240	0.3371	0.0191	0.0093	42	-0.68	036.42	0
036	Sulfur (%)	0229	0.3250	0.0300	0.3371	0.0191	0.0093	42	-0.63	036.04	0
036	Sulfur (%)	0278	0.3250	0.0100	0.3371	0.0191	0.0093	42	-0.63	036.42	0
036	Sulfur (%)	0035	0.3300	0.0200	0.3371	0.0191	0.0093	42	-0.37	036.43	0
036	Sulfur (%)	0083	0.3300	0.0000	0.3371	0.0191	0.0093	42	-0.37	036.43	0
036	Sulfur (%)	0297	0.3300	0.0000	0.3371	0.0191	0.0093	42	-0.37	036.43	0
036	Sulfur (%)	0202	0.3305	0.0030	0.3371	0.0191	0.0093	42	-0.34	036.43	0
036	Sulfur (%)	0407	0.3327	0.0078	0.3371	0.0191	0.0093	42	-0.23	036.42	0
036	Sulfur (%)	0553	0.3340	0.0020	0.3371	0.0191	0.0093	42	-0.16	036.53	0
036	Sulfur (%)	0265	0.3350	0.0100	0.3371	0.0191	0.0093	42	-0.11	036.42	0
036	Sulfur (%)	0353	0.3350	0.0100	0.3371	0.0191	0.0093	42	-0.11	036.43	0
036	Sulfur (%)	0186	0.3386	0.0010	0.3371	0.0191	0.0093	42	0.08	036.52	0
036	Sulfur (%)	0171	0.3400	0.0000	0.3371	0.0191	0.0093	42	0.15	036.42	0
036	Sulfur (%)	0726	0.3400	0.0000	0.3371	0.0191	0.0093	42	0.15	036.42	0
036	Sulfur (%)	0910	0.3411	0.0116	0.3371	0.0191	0.0093	42	0.21	036.52	0
036	Sulfur (%)	0407	0.3432	0.0224	0.3371	0.0191	0.0093	42	0.32	036.43	0
036	Sulfur (%)	0035	0.3450	0.0300	0.3371	0.0191	0.0093	42	0.41	036.42	0
036	Sulfur (%)	0164	0.3450	0.0100	0.3371	0.0191	0.0093	42	0.41	036.42	0
036	Sulfur (%)	0169	0.3450	0.0100	0.3371	0.0191	0.0093	42	0.41	036.43	0
036	Sulfur (%)	0918	0.3450	0.0060	0.3371	0.0191	0.0093	42	0.41	036.43	0
036	Sulfur (%)	0345	0.3465	0.0070	0.3371	0.0191	0.0093	42	0.49	036.43	0
036	Sulfur (%)	0560	0.3506	0.0505	0.3371	0.0191	0.0093	42	0.70	036.52	0
036	Sulfur (%)	0032	0.3507	0.0051	0.3371	0.0191	0.0093	42	0.71	036.43	0
036	Sulfur (%)	0708	0.3515	0.0030	0.3371	0.0191	0.0093	42	0.75	036.42	0
036	Sulfur (%)	0148	0.3550	0.0080	0.3371	0.0191	0.0093	42	0.94	036.04	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
036	Sulfur (%)	0190	0.3550	0.0100	0.3371	0.0191	0.0093	42	0.94	036.42	0
036	Sulfur (%)	0045	0.3560	0.0060	0.3371	0.0191	0.0093	42	0.99	036.42	0
036	Sulfur (%)	0037	0.3630	0.0080	0.3371	0.0191	0.0093	42	1.35	036.42	0
036	Sulfur (%)	2089	0.3650	0.0100	0.3371	0.0191	0.0093	42	1.46	036.43	0
036	Sulfur (%)	0510	0.3700	0.0000	0.3371	0.0191	0.0093	42	1.72	036.43	0
036	Sulfur (%)	0042	0.3810	0.0260	0.3371	0.0191	0.0093	42	2.30	036.42	0
036	Sulfur (%)	0098	0.3822	0.0036	0.3371	0.0191	0.0093	42	2.36	036.43	0
036	Sulfur (%)	0229	0.2900	0.0600	0.3371	0.0191	0.0093	42	-2.46	036.42	1
037	Zinc (ppm)	2234	235.8	18.05	448.4	31.73	15.66	85	-6.70	037.43	0
037	Zinc (ppm)	0294	372.8	16.08	448.4	31.73	15.66	85	-2.38	037.42	0
037	Zinc (ppm)	2012	387.2	2.400	448.4	31.73	15.66	85	-1.93	037.41	0
037	Zinc (ppm)	0199	398.4	4.740	448.4	31.73	15.66	85	-1.58	037.53	0
037	Zinc (ppm)	0964	402.4	25.36	448.4	31.73	15.66	85	-1.45	037.43	0
037	Zinc (ppm)	0948	404.0	3.976	448.4	31.73	15.66	85	-1.40	037.33	0
037	Zinc (ppm)	0009	405.0	20.00	448.4	31.73	15.66	85	-1.37	037.42	0
037	Zinc (ppm)	0278	407.8	15.00	448.4	31.73	15.66	85	-1.28	037.42	0
037	Zinc (ppm)	0692	411.0	38.00	448.4	31.73	15.66	85	-1.18	037.42	0
037	Zinc (ppm)	0354	414.8	1.330	448.4	31.73	15.66	85	-1.06	037.41	0
037	Zinc (ppm)	0910	418.0	7.000	448.4	31.73	15.66	85	-0.96	037.41	0
037	Zinc (ppm)	2188	418.6	3.140	448.4	31.73	15.66	85	-0.94	037.31	0
037	Zinc (ppm)	2062	419.5	20.90	448.4	31.73	15.66	85	-0.91	037.31	0
037	Zinc (ppm)	0074	422.0	18.00	448.4	31.73	15.66	85	-0.83	037.41	0
037	Zinc (ppm)	0001	422.5	4.500	448.4	31.73	15.66	85	-0.82	037.31	0
037	Zinc (ppm)	0096	424.1	3.841	448.4	31.73	15.66	85	-0.77	037.42	0
037	Zinc (ppm)	2089	424.5	15.04	448.4	31.73	15.66	85	-0.75	037.43	0
037	Zinc (ppm)	0175	427.0	26.00	448.4	31.73	15.66	85	-0.67	037.31	0
037	Zinc (ppm)	0265	428.0	4.000	448.4	31.73	15.66	85	-0.64	037.42	0
037	Zinc (ppm)	2129	428.2	14.10	448.4	31.73	15.66	85	-0.64	037.42	0
037	Zinc (ppm)	0045	428.5	21.00	448.4	31.73	15.66	85	-0.63	037.42	0
037	Zinc (ppm)	0918	428.9	20.44	448.4	31.73	15.66	85	-0.61	037.43	0
037	Zinc (ppm)	0870	429.9	12.30	448.4	31.73	15.66	85	-0.58	037.42	0
037	Zinc (ppm)	0868	430.5	15.00	448.4	31.73	15.66	85	-0.56	037.31	0
037	Zinc (ppm)	0598	430.6	10.80	448.4	31.73	15.66	85	-0.56	037.41	0
037	Zinc (ppm)	0003	431.5	11.00	448.4	31.73	15.66	85	-0.53	037.41	0
037	Zinc (ppm)	0226	433.4	8.820	448.4	31.73	15.66	85	-0.47	037.41	0
037	Zinc (ppm)	0693	433.4	72.88	448.4	31.73	15.66	85	-0.47	037.42	0
037	Zinc (ppm)	0148	433.5	21.00	448.4	31.73	15.66	85	-0.47	037.41	0
037	Zinc (ppm)	0536	435.0	4.000	448.4	31.73	15.66	85	-0.42	037.31	0
037	Zinc (ppm)	2146	435.2	13.30	448.4	31.73	15.66	85	-0.42	037.51	0
037	Zinc (ppm)	0186	436.5	3.000	448.4	31.73	15.66	85	-0.37	037.42	0
037	Zinc (ppm)	0263	437.3	6.440	448.4	31.73	15.66	85	-0.35	037.42	0
037	Zinc (ppm)	0675	438.8	41.85	448.4	31.73	15.66	85	-0.30	037.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
037	Zinc (ppm)	0560	439.2	62.50	448.4	31.73	15.66	85	-0.29	037.52	0
037	Zinc (ppm)	0939	439.5	13.00	448.4	31.73	15.66	85	-0.28	037.53	0
037	Zinc (ppm)	0529	439.7	5.300	448.4	31.73	15.66	85	-0.28	037.31	0
037	Zinc (ppm)	0590	440.0	8.000	448.4	31.73	15.66	85	-0.26	037.99	0
037	Zinc (ppm)	0510	440.5	1.000	448.4	31.73	15.66	85	-0.25	037.43	0
037	Zinc (ppm)	0171	442.0	4.900	448.4	31.73	15.66	85	-0.20	037.41	0
037	Zinc (ppm)	0345	443.0	2.000	448.4	31.73	15.66	85	-0.17	037.43	0
037	Zinc (ppm)	0164	443.0	42.00	448.4	31.73	15.66	85	-0.17	037.44	0
037	Zinc (ppm)	0019	444.9	7.100	448.4	31.73	15.66	85	-0.11	037.41	0
037	Zinc (ppm)	0941	446.0	0.0000	448.4	31.73	15.66	85	-0.08	037.43	0
037	Zinc (ppm)	2196	446.2	0.0000	448.4	31.73	15.66	85	-0.07	037.31	0
037	Zinc (ppm)	0357	446.5	1.113	448.4	31.73	15.66	85	-0.06	037.42	0
037	Zinc (ppm)	0848	448.0	0.2000	448.4	31.73	15.66	85	-0.01	037.41	0
037	Zinc (ppm)	0083	448.0	6.000	448.4	31.73	15.66	85	-0.01	037.43	0
037	Zinc (ppm)	2146	448.2	1.990	448.4	31.73	15.66	85	-0.01	037.31	0
037	Zinc (ppm)	0968	449.0	2.000	448.4	31.73	15.66	85	0.02	037.43	0
037	Zinc (ppm)	0229	450.5	24.47	448.4	31.73	15.66	85	0.06	037.41	0
037	Zinc (ppm)	0358	450.9	14.00	448.4	31.73	15.66	85	0.08	037.41	0
037	Zinc (ppm)	0186	452.5	3.000	448.4	31.73	15.66	85	0.13	037.52	0
037	Zinc (ppm)	0878	453.5	15.00	448.4	31.73	15.66	85	0.16	037.41	0
037	Zinc (ppm)	2302	453.5	13.00	448.4	31.73	15.66	85	0.16	037.99	0
037	Zinc (ppm)	0066	455.4	2.590	448.4	31.73	15.66	85	0.22	037.44	0
037	Zinc (ppm)	0202	457.7	15.46	448.4	31.73	15.66	85	0.29	037.43	0
037	Zinc (ppm)	0010	460.0	0.0000	448.4	31.73	15.66	85	0.37	037.33	0
037	Zinc (ppm)	0098	460.5	2.300	448.4	31.73	15.66	85	0.38	037.41	0
037	Zinc (ppm)	0921	461.5	50.10	448.4	31.73	15.66	85	0.41	037.31	0
037	Zinc (ppm)	0042	461.5	77.00	448.4	31.73	15.66	85	0.41	037.42	0
037	Zinc (ppm)	0169	461.5	3.000	448.4	31.73	15.66	85	0.41	037.43	0
037	Zinc (ppm)	0208	466.5	57.00	448.4	31.73	15.66	85	0.57	037.31	0
037	Zinc (ppm)	0035	468.0	14.00	448.4	31.73	15.66	85	0.62	037.43	0
037	Zinc (ppm)	0553	477.5	13.00	448.4	31.73	15.66	85	0.92	037.53	0
037	Zinc (ppm)	0682	478.0	0.0000	448.4	31.73	15.66	85	0.93	037.43	0
037	Zinc (ppm)	0407	478.8	4.487	448.4	31.73	15.66	85	0.96	037.41	0
037	Zinc (ppm)	0208	479.6	0.1000	448.4	31.73	15.66	85	0.98	037.41	0
037	Zinc (ppm)	0505	480.5	34.30	448.4	31.73	15.66	85	1.01	037.43	0
037	Zinc (ppm)	0096	480.5	17.27	448.4	31.73	15.66	85	1.01	037.52	0
037	Zinc (ppm)	0726	480.9	1.500	448.4	31.73	15.66	85	1.02	037.42	0
037	Zinc (ppm)	0425	485.0	6.300	448.4	31.73	15.66	85	1.15	037.43	0
037	Zinc (ppm)	0939	485.5	12.87	448.4	31.73	15.66	85	1.17	037.31	0
037	Zinc (ppm)	0032	487.2	7.500	448.4	31.73	15.66	85	1.22	037.43	0
037	Zinc (ppm)	0353	488.5	23.00	448.4	31.73	15.66	85	1.26	037.43	0
037	Zinc (ppm)	0190	490.2	1.170	448.4	31.73	15.66	85	1.32	037.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
037	Zinc (ppm)	0297	493.5	21.00	448.4	31.73	15.66	85	1.42	037.43	0
037	Zinc (ppm)	2178	500.5	11.00	448.4	31.73	15.66	85	1.64	037.33	0
037	Zinc (ppm)	0572	509.5	77.00	448.4	31.73	15.66	85	1.93	037.53	0
037	Zinc (ppm)	0035	513.5	7.000	448.4	31.73	15.66	85	2.05	037.42	0
037	Zinc (ppm)	0100	513.5	39.00	448.4	31.73	15.66	85	2.05	037.99	0
037	Zinc (ppm)	0656	518.0	0.0000	448.4	31.73	15.66	85	2.19	037.31	0
037	Zinc (ppm)	0407	522.6	23.48	448.4	31.73	15.66	85	2.34	037.43	0
037	Zinc (ppm)	0098	528.4	48.90	448.4	31.73	15.66	85	2.52	037.43	0
037	Zinc (ppm)	0037	543.4	10.50	448.4	31.73	15.66	85	2.99	037.42	0
037	Zinc (ppm)	0520	522.5	119.0	448.4	31.73	15.66	85	2.34	037.41	1
037	Zinc (ppm)	0884	912.9	3.000	448.4	31.73	15.66	85	14.64	037.31	2
038	Molybdenum (ppm)	0171	2.270	0.0200	2.989	0.3336	0.1300	22	-2.16	038.41	0
038	Molybdenum (ppm)	0278	2.520	0.1400	2.989	0.3336	0.1300	22	-1.41	038.42	0
038	Molybdenum (ppm)	0407	2.558	0.0425	2.989	0.3336	0.1300	22	-1.29	038.41	0
038	Molybdenum (ppm)	0297	2.565	0.5100	2.989	0.3336	0.1300	22	-1.27	038.43	0
038	Molybdenum (ppm)	0345	2.650	0.0200	2.989	0.3336	0.1300	22	-1.02	038.43	0
038	Molybdenum (ppm)	0353	2.754	0.3005	2.989	0.3336	0.1300	22	-0.71	038.43	0
038	Molybdenum (ppm)	0226	2.870	0.0200	2.989	0.3336	0.1300	22	-0.36	038.41	0
038	Molybdenum (ppm)	0910	2.925	0.1300	2.989	0.3336	0.1300	22	-0.19	038.52	0
038	Molybdenum (ppm)	0186	2.995	0.0100	2.989	0.3336	0.1300	22	0.02	038.52	0
038	Molybdenum (ppm)	0939	3.010	0.0000	2.989	0.3336	0.1300	22	0.06	038.53	0
038	Molybdenum (ppm)	0693	3.036	0.1900	2.989	0.3336	0.1300	22	0.14	038.42	0
038	Molybdenum (ppm)	0083	3.100	0.0000	2.989	0.3336	0.1300	22	0.33	038.43	0
038	Molybdenum (ppm)	0202	3.105	0.1700	2.989	0.3336	0.1300	22	0.35	038.43	0
038	Molybdenum (ppm)	0164	3.105	0.2700	2.989	0.3336	0.1300	22	0.35	038.53	0
038	Molybdenum (ppm)	0169	3.110	0.1800	2.989	0.3336	0.1300	22	0.36	038.43	0
038	Molybdenum (ppm)	0918	3.116	0.1524	2.989	0.3336	0.1300	22	0.38	038.53	0
038	Molybdenum (ppm)	0096	3.163	0.2170	2.989	0.3336	0.1300	22	0.52	038.42	0
038	Molybdenum (ppm)	0407	3.189	0.2369	2.989	0.3336	0.1300	22	0.60	038.43	0
038	Molybdenum (ppm)	0096	3.225	0.1110	2.989	0.3336	0.1300	22	0.71	038.52	0
038	Molybdenum (ppm)	0553	3.295	0.0100	2.989	0.3336	0.1300	22	0.92	038.53	0
038	Molybdenum (ppm)	0510	3.550	0.1000	2.989	0.3336	0.1300	22	1.68	038.43	0
038	Molybdenum (ppm)	0045	4.275	0.0300	2.989	0.3336	0.1300	22	3.86	038.42	0
038	Molybdenum (ppm)	0265	4.380	0.8200	2.989	0.3336	0.1300	22	4.17	038.42	1
040	Barium (ppm)	0560	4.215	0.5700				3		040.52	0
040	Barium (ppm)	0096	4.425	0.3080				3		040.52	0
040	Barium (ppm)	0918	4.450	0.2816				3		040.53	0
041	Vanadium (ppm)	0553	2.035	0.0100				1		041.53	0
042	Chloride (%)	0948	0.9260	0.0060				3		042.00	0
042	Chloride (%)	0297	0.9650	0.0700				3		042.99	0
042	Chloride (%)	0066	1.030	0.0470				3		042.99	0
101	Choline Chloride (ppm)	0227	1,545	70.00				1		101.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
102	Niacin (ppm)	0227	34.15	2.900				1		102.01	0
103	Pantothenic Acid (ppm)	0227	11.90	0.0000				1		103.01	0
104	Riboflavin (ppm)	0910	2.060	0.0600				2		104.03	0
104	Riboflavin (ppm)	0227	6.500	0.5200				2		104.00	0
105	Thiamine (ppm)	0227	58.85	4.100				2		105.01	0
105	Thiamine (ppm)	0910	67.30	3.680				2		105.00	0
106	Vitamin A (KU / kg)	0199	18.59	8.520	30.30	7.911	4.559	15	-1.48	106.02	0
106	Vitamin A (KU / kg)	2053	22.30	2.800	30.30	7.911	4.559	15	-1.01	106.02	0
106	Vitamin A (KU / kg)	0941	22.45	7.825	30.30	7.911	4.559	15	-0.99	106.02	0
106	Vitamin A (KU / kg)	0035	25.06	3.780	30.30	7.911	4.559	15	-0.66	106.02	0
106	Vitamin A (KU / kg)	0098	26.50	18.60	30.30	7.911	4.559	15	-0.48	106.01	0
106	Vitamin A (KU / kg)	0148	28.04	0.0410	30.30	7.911	4.559	15	-0.29	106.02	0
106	Vitamin A (KU / kg)	0910	28.20	0.2000	30.30	7.911	4.559	15	-0.27	106.02	0
106	Vitamin A (KU / kg)	0227	29.55	8.300	30.30	7.911	4.559	15	-0.10	106.00	0
106	Vitamin A (KU / kg)	0675	29.92	0.6700	30.30	7.911	4.559	15	-0.05	106.02	0
106	Vitamin A (KU / kg)	0010	32.55	3.300	30.30	7.911	4.559	15	0.28	106.02	0
106	Vitamin A (KU / kg)	0036	33.20	5.400	30.30	7.911	4.559	15	0.37	106.02	0
106	Vitamin A (KU / kg)	2152	37.63	1.120	30.30	7.911	4.559	15	0.93	106.02	0
106	Vitamin A (KU / kg)	0964	38.69	0.6614	30.30	7.911	4.559	15	1.06	106.00	0
106	Vitamin A (KU / kg)	0208	39.68	6.173	30.30	7.911	4.559	15	1.19	106.02	0
106	Vitamin A (KU / kg)	0169	44.50	1.000	30.30	7.911	4.559	15	1.79	106.02	0
107	Vitamin B12 (ppb)	0227	11.90	1.000				1		107.00	0
108	Vitamin D3 (KU / kg)	0914	2.860	0.1200	3.388	0.4233	0.2400	4		108.02	0
108	Vitamin D3 (KU / kg)	0227	3.290	0.4200	3.388	0.4233	0.2400	4		108.02	0
108	Vitamin D3 (KU / kg)	0675	3.535	0.3700	3.388	0.4233	0.2400	4		108.02	0
108	Vitamin D3 (KU / kg)	0910	3.865	0.0500	3.388	0.4233	0.2400	4		108.02	0
108	Vitamin D3 (KU / kg)	0208	3.105	2.050	3.388	0.4233	0.2400	4		108.02	1
109	Vitamin E (IU / kg)	0675	60.25	0.7400	112.5	21.93	4.217	16	-2.38	109.02	0
109	Vitamin E (IU / kg)	0921	77.41	0.7600	112.5	21.93	4.217	16	-1.60	109.02	0
109	Vitamin E (IU / kg)	0169	82.50	1.000	112.5	21.93	4.217	16	-1.37	109.02	0
109	Vitamin E (IU / kg)	0148	99.56	9.000	112.5	21.93	4.217	16	-0.59	109.02	0
109	Vitamin E (IU / kg)	0208	108.9	8.199	112.5	21.93	4.217	16	-0.17	109.02	0
109	Vitamin E (IU / kg)	2152	110.7	0.1600	112.5	21.93	4.217	16	-0.08	109.02	0
109	Vitamin E (IU / kg)	2103	111.5	5.000	112.5	21.93	4.217	16	-0.05	109.02	0
109	Vitamin E (IU / kg)	0098	112.5	1.000	112.5	21.93	4.217	16	0.00	109.02	0
109	Vitamin E (IU / kg)	0160	116.5	9.000	112.5	21.93	4.217	16	0.18	109.99	0
109	Vitamin E (IU / kg)	0910	117.5	3.000	112.5	21.93	4.217	16	0.23	109.02	0
109	Vitamin E (IU / kg)	0171	119.0	6.000	112.5	21.93	4.217	16	0.29	109.99	0
109	Vitamin E (IU / kg)	0590	125.0	2.000	112.5	21.93	4.217	16	0.57	109.02	0
109	Vitamin E (IU / kg)	0227	127.0	2.000	112.5	21.93	4.217	16	0.66	109.02	0
109	Vitamin E (IU / kg)	0036	129.8	3.600	112.5	21.93	4.217	16	0.79	109.02	0
109	Vitamin E (IU / kg)	0941	135.6	11.69	112.5	21.93	4.217	16	1.05	109.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
109	Vitamin E (IU / kg)	0199	158.9	4.330	112.5	21.93	4.217	16	2.12	109.02	0
111	Vitamin C, Phosphorylated (ppm)	0227	< 4.4					0		111.00	5
112	Pyridoxine (µg / g)	0227	2.760	0.0200				1		112.01	0
113	Folic Acid (ppm)	0227	1.360	0.0800				1		113.01	0
114	Biotin (ppm)	0227	1.225	0.0100				1		114.01	0
115	Non Protein N (NPN) (%)	0098	1.140	0.0200				1		115.00	0
118	Peroxide value (meq/kg)	2081	2.630	0.1600				1		118.99	0
120	Alanine (%)	0918	0.6996	0.0287	0.7705	0.0278	0.0140	27	-2.55	120.00	0
120	Alanine (%)	0353	0.7150	0.0100	0.7705	0.0278	0.0140	27	-2.00	120.00	0
120	Alanine (%)	0626	0.7365	0.0030	0.7705	0.0278	0.0140	27	-1.22	120.05	0
120	Alanine (%)	0682	0.7380	0.0000	0.7705	0.0278	0.0140	27	-1.17	120.00	0
120	Alanine (%)	0910	0.7500	0.0000	0.7705	0.0278	0.0140	27	-0.74	120.00	0
120	Alanine (%)	0148	0.7500	0.0200	0.7705	0.0278	0.0140	27	-0.74	120.05	0
120	Alanine (%)	0676	0.7500	0.0000	0.7705	0.0278	0.0140	27	-0.74	120.05	0
120	Alanine (%)	0354	0.7540	0.0260	0.7705	0.0278	0.0140	27	-0.59	120.00	0
120	Alanine (%)	0859	0.7542	0.0118	0.7705	0.0278	0.0140	27	-0.59	120.00	0
120	Alanine (%)	0889	0.7550	0.0500	0.7705	0.0278	0.0140	27	-0.56	120.99	0
120	Alanine (%)	0868	0.7585	0.0290	0.7705	0.0278	0.0140	27	-0.43	120.00	0
120	Alanine (%)	0407	0.7610	0.0200	0.7705	0.0278	0.0140	27	-0.34	120.05	0
120	Alanine (%)	0227	0.7700	0.0400	0.7705	0.0278	0.0140	27	-0.02	120.99	0
120	Alanine (%)	0652	0.7750	0.0100	0.7705	0.0278	0.0140	27	0.16	120.00	0
120	Alanine (%)	0968	0.7750	0.0100	0.7705	0.0278	0.0140	27	0.16	120.00	0
120	Alanine (%)	0098	0.7780	0.0040	0.7705	0.0278	0.0140	27	0.27	120.02	0
120	Alanine (%)	0675	0.7800	0.0400	0.7705	0.0278	0.0140	27	0.34	120.00	0
120	Alanine (%)	0226	0.7826	0.0095	0.7705	0.0278	0.0140	27	0.44	120.00	0
120	Alanine (%)	0208	0.7850	0.0100	0.7705	0.0278	0.0140	27	0.52	120.00	0
120	Alanine (%)	0571	0.7865	0.0030	0.7705	0.0278	0.0140	27	0.58	120.00	0
120	Alanine (%)	0878	0.7910	0.0020	0.7705	0.0278	0.0140	27	0.74	120.00	0
120	Alanine (%)	0870	0.7957	0.0073	0.7705	0.0278	0.0140	27	0.91	120.00	0
120	Alanine (%)	2059	0.7965	0.0030	0.7705	0.0278	0.0140	27	0.94	120.00	0
120	Alanine (%)	0941	0.7985	0.0370	0.7705	0.0278	0.0140	27	1.01	120.00	0
120	Alanine (%)	0872	0.7995	0.0050	0.7705	0.0278	0.0140	27	1.05	120.00	0
120	Alanine (%)	2196	0.8280	0.0000	0.7705	0.0278	0.0140	27	2.07	120.05	0
120	Alanine (%)	2188	0.9960	0.0000	0.7705	0.0278	0.0140	27	8.12	120.05	0
121	Arginine (%)	0859	0.8819	0.0232	0.9778	0.0465	0.0220	26	-2.06	121.00	0
121	Arginine (%)	0682	0.9110	0.0000	0.9778	0.0465	0.0220	26	-1.44	121.00	0
121	Arginine (%)	0889	0.9200	0.0600	0.9778	0.0465	0.0220	26	-1.24	121.99	0
121	Arginine (%)	0918	0.9280	0.0430	0.9778	0.0465	0.0220	26	-1.07	121.00	0
121	Arginine (%)	0353	0.9350	0.0100	0.9778	0.0465	0.0220	26	-0.92	121.00	0
121	Arginine (%)	0227	0.9350	0.0100	0.9778	0.0465	0.0220	26	-0.92	121.99	0
121	Arginine (%)	0626	0.9420	0.0080	0.9778	0.0465	0.0220	26	-0.77	121.05	0
121	Arginine (%)	0968	0.9550	0.0100	0.9778	0.0465	0.0220	26	-0.49	121.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
121	Arginine (%)	0652	0.9600	0.0200	0.9778	0.0465	0.0220	26	-0.38	121.00	0
121	Arginine (%)	2196	0.9610	0.0000	0.9778	0.0465	0.0220	26	-0.36	121.05	0
121	Arginine (%)	0870	0.9630	0.0200	0.9778	0.0465	0.0220	26	-0.32	121.00	0
121	Arginine (%)	0407	0.9740	0.0480	0.9778	0.0465	0.0220	26	-0.08	121.05	0
121	Arginine (%)	0354	0.9750	0.0020	0.9778	0.0465	0.0220	26	-0.06	121.00	0
121	Arginine (%)	0868	0.9795	0.0110	0.9778	0.0465	0.0220	26	0.04	121.00	0
121	Arginine (%)	0098	0.9820	0.0020	0.9778	0.0465	0.0220	26	0.09	121.02	0
121	Arginine (%)	0675	0.9950	0.0500	0.9778	0.0465	0.0220	26	0.37	121.00	0
121	Arginine (%)	0910	1.000	0.0800	0.9778	0.0465	0.0220	26	0.48	121.00	0
121	Arginine (%)	2059	1.003	0.0110	0.9778	0.0465	0.0220	26	0.53	121.00	0
121	Arginine (%)	0571	1.009	0.0160	0.9778	0.0465	0.0220	26	0.67	121.00	0
121	Arginine (%)	0878	1.013	0.0160	0.9778	0.0465	0.0220	26	0.76	121.00	0
121	Arginine (%)	0148	1.015	0.0300	0.9778	0.0465	0.0220	26	0.80	121.05	0
121	Arginine (%)	0676	1.015	0.0100	0.9778	0.0465	0.0220	26	0.80	121.05	0
121	Arginine (%)	0872	1.020	0.0220	0.9778	0.0465	0.0220	26	0.91	121.00	0
121	Arginine (%)	0941	1.038	0.0500	0.9778	0.0465	0.0220	26	1.29	121.00	0
121	Arginine (%)	0208	1.040	0.0200	0.9778	0.0465	0.0220	26	1.34	121.00	0
121	Arginine (%)	2188	1.076	0.0000	0.9778	0.0465	0.0220	26	2.11	121.05	0
121	Arginine (%)	0226	1.010	0.1402	0.9778	0.0465	0.0220	26	0.70	121.00	1
122	Aspartic (%)	0889	1.355	0.0300	1.615	0.0582	0.0317	27	-4.47	122.99	0
122	Aspartic (%)	0859	1.452	0.0344	1.615	0.0582	0.0317	27	-2.81	122.00	0
122	Aspartic (%)	0682	1.479	0.0000	1.615	0.0582	0.0317	27	-2.34	122.00	0
122	Aspartic (%)	0148	1.530	0.0600	1.615	0.0582	0.0317	27	-1.47	122.05	0
122	Aspartic (%)	0652	1.540	0.0000	1.615	0.0582	0.0317	27	-1.30	122.00	0
122	Aspartic (%)	2188	1.570	0.0000	1.615	0.0582	0.0317	27	-0.78	122.05	0
122	Aspartic (%)	0407	1.574	0.0470	1.615	0.0582	0.0317	27	-0.72	122.05	0
122	Aspartic (%)	0626	1.592	0.0260	1.615	0.0582	0.0317	27	-0.40	122.05	0
122	Aspartic (%)	0968	1.600	0.0000	1.615	0.0582	0.0317	27	-0.27	122.00	0
122	Aspartic (%)	0354	1.603	0.0280	1.615	0.0582	0.0317	27	-0.21	122.00	0
122	Aspartic (%)	0208	1.615	0.0300	1.615	0.0582	0.0317	27	-0.01	122.00	0
122	Aspartic (%)	0910	1.615	0.0900	1.615	0.0582	0.0317	27	-0.01	122.00	0
122	Aspartic (%)	0676	1.620	0.0400	1.615	0.0582	0.0317	27	0.08	122.05	0
122	Aspartic (%)	0918	1.623	0.1059	1.615	0.0582	0.0317	27	0.12	122.00	0
122	Aspartic (%)	0571	1.631	0.0140	1.615	0.0582	0.0317	27	0.27	122.00	0
122	Aspartic (%)	0353	1.635	0.0100	1.615	0.0582	0.0317	27	0.34	122.00	0
122	Aspartic (%)	0227	1.635	0.0300	1.615	0.0582	0.0317	27	0.34	122.99	0
122	Aspartic (%)	0941	1.640	0.1290	1.615	0.0582	0.0317	27	0.41	122.00	0
122	Aspartic (%)	0675	1.650	0.0200	1.615	0.0582	0.0317	27	0.59	122.00	0
122	Aspartic (%)	0098	1.657	0.0160	1.615	0.0582	0.0317	27	0.71	122.02	0
122	Aspartic (%)	0870	1.660	0.0255	1.615	0.0582	0.0317	27	0.76	122.00	0
122	Aspartic (%)	2196	1.666	0.0000	1.615	0.0582	0.0317	27	0.87	122.05	0
122	Aspartic (%)	0878	1.669	0.0070	1.615	0.0582	0.0317	27	0.91	122.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
122	Aspartic (%)	2059	1.670	0.0170	1.615	0.0582	0.0317	27	0.93	122.00	0
122	Aspartic (%)	0226	1.670	0.0860	1.615	0.0582	0.0317	27	0.93	122.00	0
122	Aspartic (%)	0872	1.675	0.0080	1.615	0.0582	0.0317	27	1.02	122.00	0
122	Aspartic (%)	0868	1.697	0.0010	1.615	0.0582	0.0317	27	1.39	122.00	0
124	Cysteine/Cystine (%)	0870	0.2127	0.0141	0.2701	0.0250	0.0076	24	-2.30	124.00	0
124	Cysteine/Cystine (%)	0889	0.2200	0.0000	0.2701	0.0250	0.0076	24	-2.00	124.99	0
124	Cysteine/Cystine (%)	0859	0.2351	0.0036	0.2701	0.0250	0.0076	24	-1.40	124.00	0
124	Cysteine/Cystine (%)	2188	0.2400	0.0000	0.2701	0.0250	0.0076	24	-1.20	124.05	0
124	Cysteine/Cystine (%)	0652	0.2500	0.0000	0.2701	0.0250	0.0076	24	-0.81	124.00	0
124	Cysteine/Cystine (%)	0910	0.2500	0.0000	0.2701	0.0250	0.0076	24	-0.81	124.00	0
124	Cysteine/Cystine (%)	0226	0.2641	0.0359	0.2701	0.0250	0.0076	24	-0.24	124.00	0
124	Cysteine/Cystine (%)	0353	0.2650	0.0100	0.2701	0.0250	0.0076	24	-0.21	124.00	0
124	Cysteine/Cystine (%)	0878	0.2650	0.0000	0.2701	0.0250	0.0076	24	-0.21	124.00	0
124	Cysteine/Cystine (%)	0676	0.2700	0.0000	0.2701	0.0250	0.0076	24	-0.01	124.05	0
124	Cysteine/Cystine (%)	0148	0.2705	0.0450	0.2701	0.0250	0.0076	24	0.01	124.05	0
124	Cysteine/Cystine (%)	0571	0.2715	0.0030	0.2701	0.0250	0.0076	24	0.05	124.00	0
124	Cysteine/Cystine (%)	0968	0.2715	0.0030	0.2701	0.0250	0.0076	24	0.05	124.00	0
124	Cysteine/Cystine (%)	0868	0.2735	0.0010	0.2701	0.0250	0.0076	24	0.13	124.00	0
124	Cysteine/Cystine (%)	2059	0.2735	0.0010	0.2701	0.0250	0.0076	24	0.13	124.00	0
124	Cysteine/Cystine (%)	0227	0.2750	0.0100	0.2701	0.0250	0.0076	24	0.19	124.99	0
124	Cysteine/Cystine (%)	0098	0.2795	0.0010	0.2701	0.0250	0.0076	24	0.37	124.02	0
124	Cysteine/Cystine (%)	0941	0.2815	0.0090	0.2701	0.0250	0.0076	24	0.45	124.00	0
124	Cysteine/Cystine (%)	0872	0.2835	0.0090	0.2701	0.0250	0.0076	24	0.53	124.00	0
124	Cysteine/Cystine (%)	0407	0.2841	0.0061	0.2701	0.0250	0.0076	24	0.56	124.05	0
124	Cysteine/Cystine (%)	2196	0.2920	0.0000	0.2701	0.0250	0.0076	24	0.87	124.05	0
124	Cysteine/Cystine (%)	0354	0.3285	0.0110	0.2701	0.0250	0.0076	24	2.33	124.00	0
124	Cysteine/Cystine (%)	0675	0.3600	0.0200	0.2701	0.0250	0.0076	24	3.59	124.00	0
124	Cysteine/Cystine (%)	0682	0.3870	0.0000	0.2701	0.0250	0.0076	24	4.67	124.00	0
124	Cysteine/Cystine (%)	0918	0.3050	0.0700	0.2701	0.0250	0.0076	24	1.39	124.00	1
124	Cysteine/Cystine (%)	2319	< 0.005		0.2701	0.0250	0.0076	24		124.05	5
125	Glutamic (%)	0889	2.230	0.0400	2.833	0.1497	0.0497	27	-4.02	125.99	0
125	Glutamic (%)	0676	2.555	0.0100	2.833	0.1497	0.0497	27	-1.85	125.05	0
125	Glutamic (%)	0682	2.573	0.0000	2.833	0.1497	0.0497	27	-1.73	125.00	0
125	Glutamic (%)	0859	2.594	0.0553	2.833	0.1497	0.0497	27	-1.59	125.00	0
125	Glutamic (%)	0407	2.703	0.1090	2.833	0.1497	0.0497	27	-0.87	125.05	0
125	Glutamic (%)	0354	2.719	0.0060	2.833	0.1497	0.0497	27	-0.76	125.00	0
125	Glutamic (%)	0968	2.770	0.0400	2.833	0.1497	0.0497	27	-0.42	125.00	0
125	Glutamic (%)	0227	2.785	0.0300	2.833	0.1497	0.0497	27	-0.32	125.99	0
125	Glutamic (%)	0652	2.790	0.0800	2.833	0.1497	0.0497	27	-0.28	125.00	0
125	Glutamic (%)	0626	2.794	0.0260	2.833	0.1497	0.0497	27	-0.26	125.05	0
125	Glutamic (%)	2188	2.807	0.0000	2.833	0.1497	0.0497	27	-0.17	125.05	0
125	Glutamic (%)	0571	2.809	0.0380	2.833	0.1497	0.0497	27	-0.16	125.00	0



Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
125	Glutamic (%)	2059	2.825	0.0260	2.833	0.1497	0.0497	27	-0.05	125.00	0
125	Glutamic (%)	0868	2.836	0.0110	2.833	0.1497	0.0497	27	0.02	125.00	0
125	Glutamic (%)	0872	2.842	0.0160	2.833	0.1497	0.0497	27	0.06	125.00	0
125	Glutamic (%)	0910	2.850	0.1400	2.833	0.1497	0.0497	27	0.12	125.00	0
125	Glutamic (%)	0918	2.874	0.0247	2.833	0.1497	0.0497	27	0.28	125.00	0
125	Glutamic (%)	0941	2.886	0.1440	2.833	0.1497	0.0497	27	0.36	125.00	0
125	Glutamic (%)	0098	2.889	0.0180	2.833	0.1497	0.0497	27	0.38	125.02	0
125	Glutamic (%)	2196	2.905	0.0000	2.833	0.1497	0.0497	27	0.48	125.05	0
125	Glutamic (%)	0353	2.945	0.0100	2.833	0.1497	0.0497	27	0.75	125.00	0
125	Glutamic (%)	0878	2.950	0.0100	2.833	0.1497	0.0497	27	0.78	125.00	0
125	Glutamic (%)	0675	2.970	0.1600	2.833	0.1497	0.0497	27	0.92	125.00	0
125	Glutamic (%)	0226	2.988	0.0912	2.833	0.1497	0.0497	27	1.04	125.00	0
125	Glutamic (%)	0870	2.997	0.0060	2.833	0.1497	0.0497	27	1.10	125.00	0
125	Glutamic (%)	0148	3.075	0.1900	2.833	0.1497	0.0497	27	1.62	125.05	0
125	Glutamic (%)	0208	3.090	0.0600	2.833	0.1497	0.0497	27	1.72	125.00	0
126	Glycine (%)	0889	0.3900	0.0000	0.6973	0.0276	0.0107	27	-11.15	126.99	0
126	Glycine (%)	0682	0.6300	0.0000	0.6973	0.0276	0.0107	27	-2.44	126.00	0
126	Glycine (%)	0859	0.6566	0.0138	0.6973	0.0276	0.0107	27	-1.48	126.00	0
126	Glycine (%)	0918	0.6609	0.0202	0.6973	0.0276	0.0107	27	-1.32	126.00	0
126	Glycine (%)	0353	0.6750	0.0100	0.6973	0.0276	0.0107	27	-0.81	126.00	0
126	Glycine (%)	0910	0.6800	0.0200	0.6973	0.0276	0.0107	27	-0.63	126.00	0
126	Glycine (%)	0968	0.6800	0.0000	0.6973	0.0276	0.0107	27	-0.63	126.00	0
126	Glycine (%)	0354	0.6815	0.0030	0.6973	0.0276	0.0107	27	-0.57	126.00	0
126	Glycine (%)	0868	0.6815	0.0050	0.6973	0.0276	0.0107	27	-0.57	126.00	0
126	Glycine (%)	0870	0.6900	0.0042	0.6973	0.0276	0.0107	27	-0.26	126.00	0
126	Glycine (%)	0227	0.6900	0.0200	0.6973	0.0276	0.0107	27	-0.26	126.99	0
126	Glycine (%)	0208	0.6950	0.0100	0.6973	0.0276	0.0107	27	-0.08	126.00	0
126	Glycine (%)	0676	0.6950	0.0100	0.6973	0.0276	0.0107	27	-0.08	126.05	0
126	Glycine (%)	0652	0.7000	0.0200	0.6973	0.0276	0.0107	27	0.10	126.00	0
126	Glycine (%)	2196	0.7000	0.0000	0.6973	0.0276	0.0107	27	0.10	126.05	0
126	Glycine (%)	0571	0.7010	0.0060	0.6973	0.0276	0.0107	27	0.14	126.00	0
126	Glycine (%)	0878	0.7050	0.0000	0.6973	0.0276	0.0107	27	0.28	126.00	0
126	Glycine (%)	0098	0.7085	0.0030	0.6973	0.0276	0.0107	27	0.41	126.02	0
126	Glycine (%)	0675	0.7100	0.0400	0.6973	0.0276	0.0107	27	0.46	126.00	0
126	Glycine (%)	2188	0.7190	0.0000	0.6973	0.0276	0.0107	27	0.79	126.05	0
126	Glycine (%)	2059	0.7195	0.0010	0.6973	0.0276	0.0107	27	0.81	126.00	0
126	Glycine (%)	0626	0.7195	0.0030	0.6973	0.0276	0.0107	27	0.81	126.05	0
126	Glycine (%)	0872	0.7225	0.0050	0.6973	0.0276	0.0107	27	0.92	126.00	0
126	Glycine (%)	0941	0.7225	0.0370	0.6973	0.0276	0.0107	27	0.92	126.00	0
126	Glycine (%)	0226	0.7243	0.0053	0.6973	0.0276	0.0107	27	0.98	126.00	0
126	Glycine (%)	0148	0.7650	0.0500	0.6973	0.0276	0.0107	27	2.46	126.05	0
126	Glycine (%)	0407	0.8030	0.0020	0.6973	0.0276	0.0107	27	3.84	126.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
127	Histidine (%)	0859	0.3429	0.0071	0.3885	0.0178	0.0090	27	-2.56	127.00	0
127	Histidine (%)	0870	0.3553	0.0122	0.3885	0.0178	0.0090	27	-1.86	127.00	0
127	Histidine (%)	0676	0.3650	0.0300	0.3885	0.0178	0.0090	27	-1.32	127.05	0
127	Histidine (%)	0968	0.3700	0.0000	0.3885	0.0178	0.0090	27	-1.04	127.00	0
127	Histidine (%)	0353	0.3750	0.0100	0.3885	0.0178	0.0090	27	-0.76	127.00	0
127	Histidine (%)	0354	0.3770	0.0020	0.3885	0.0178	0.0090	27	-0.64	127.00	0
127	Histidine (%)	0226	0.3772	0.0155	0.3885	0.0178	0.0090	27	-0.64	127.00	0
127	Histidine (%)	0626	0.3785	0.0090	0.3885	0.0178	0.0090	27	-0.56	127.05	0
127	Histidine (%)	0682	0.3790	0.0000	0.3885	0.0178	0.0090	27	-0.53	127.00	0
127	Histidine (%)	0227	0.3800	0.0000	0.3885	0.0178	0.0090	27	-0.48	127.99	0
127	Histidine (%)	0918	0.3874	0.0096	0.3885	0.0178	0.0090	27	-0.06	127.00	0
127	Histidine (%)	2059	0.3880	0.0020	0.3885	0.0178	0.0090	27	-0.03	127.00	0
127	Histidine (%)	0098	0.3890	0.0000	0.3885	0.0178	0.0090	27	0.03	127.02	0
127	Histidine (%)	0571	0.3895	0.0010	0.3885	0.0178	0.0090	27	0.06	127.00	0
127	Histidine (%)	0208	0.3900	0.0000	0.3885	0.0178	0.0090	27	0.09	127.00	0
127	Histidine (%)	0675	0.3900	0.0400	0.3885	0.0178	0.0090	27	0.09	127.00	0
127	Histidine (%)	0910	0.3900	0.0200	0.3885	0.0178	0.0090	27	0.09	127.00	0
127	Histidine (%)	0148	0.3900	0.0400	0.3885	0.0178	0.0090	27	0.09	127.05	0
127	Histidine (%)	0889	0.3900	0.0000	0.3885	0.0178	0.0090	27	0.09	127.99	0
127	Histidine (%)	0868	0.3915	0.0070	0.3885	0.0178	0.0090	27	0.17	127.00	0
127	Histidine (%)	0872	0.4015	0.0010	0.3885	0.0178	0.0090	27	0.73	127.00	0
127	Histidine (%)	0878	0.4020	0.0040	0.3885	0.0178	0.0090	27	0.76	127.00	0
127	Histidine (%)	0941	0.4095	0.0210	0.3885	0.0178	0.0090	27	1.18	127.00	0
127	Histidine (%)	2188	0.4120	0.0000	0.3885	0.0178	0.0090	27	1.32	127.05	0
127	Histidine (%)	0407	0.4130	0.0120	0.3885	0.0178	0.0090	27	1.37	127.05	0
127	Histidine (%)	0652	0.4200	0.0000	0.3885	0.0178	0.0090	27	1.77	127.00	0
127	Histidine (%)	2196	0.4720	0.0000	0.3885	0.0178	0.0090	27	4.68	127.05	0
128	Isoleucine (%)	0889	0.4175	0.0050	0.6365	0.0522	0.0131	26	-4.20	128.99	0
128	Isoleucine (%)	0682	0.5460	0.0000	0.6365	0.0522	0.0131	26	-1.73	128.00	0
128	Isoleucine (%)	0353	0.5550	0.0100	0.6365	0.0522	0.0131	26	-1.56	128.00	0
128	Isoleucine (%)	0918	0.5712	0.0291	0.6365	0.0522	0.0131	26	-1.25	128.00	0
128	Isoleucine (%)	0676	0.5950	0.0100	0.6365	0.0522	0.0131	26	-0.79	128.05	0
128	Isoleucine (%)	0868	0.6080	0.0000	0.6365	0.0522	0.0131	26	-0.55	128.00	0
128	Isoleucine (%)	0968	0.6100	0.0000	0.6365	0.0522	0.0131	26	-0.51	128.00	0
128	Isoleucine (%)	0652	0.6150	0.0100	0.6365	0.0522	0.0131	26	-0.41	128.00	0
128	Isoleucine (%)	0227	0.6200	0.0000	0.6365	0.0522	0.0131	26	-0.32	128.99	0
128	Isoleucine (%)	0626	0.6235	0.0110	0.6365	0.0522	0.0131	26	-0.25	128.05	0
128	Isoleucine (%)	0870	0.6258	0.0097	0.6365	0.0522	0.0131	26	-0.21	128.00	0
128	Isoleucine (%)	0859	0.6272	0.0139	0.6365	0.0522	0.0131	26	-0.18	128.00	0
128	Isoleucine (%)	0354	0.6285	0.0010	0.6365	0.0522	0.0131	26	-0.15	128.00	0
128	Isoleucine (%)	0208	0.6350	0.0100	0.6365	0.0522	0.0131	26	-0.03	128.00	0
128	Isoleucine (%)	0910	0.6400	0.0400	0.6365	0.0522	0.0131	26	0.07	128.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
128	Isoleucine (%)	0098	0.6425	0.0030	0.6365	0.0522	0.0131	26	0.12	128.02	0
128	Isoleucine (%)	0878	0.6485	0.0130	0.6365	0.0522	0.0131	26	0.23	128.00	0
128	Isoleucine (%)	0571	0.6635	0.0030	0.6365	0.0522	0.0131	26	0.52	128.00	0
128	Isoleucine (%)	0407	0.6715	0.0450	0.6365	0.0522	0.0131	26	0.67	128.05	0
128	Isoleucine (%)	0941	0.6740	0.0440	0.6365	0.0522	0.0131	26	0.72	128.00	0
128	Isoleucine (%)	0872	0.6780	0.0080	0.6365	0.0522	0.0131	26	0.80	128.00	0
128	Isoleucine (%)	2059	0.6830	0.0060	0.6365	0.0522	0.0131	26	0.89	128.00	0
128	Isoleucine (%)	2196	0.6890	0.0000	0.6365	0.0522	0.0131	26	1.01	128.05	0
128	Isoleucine (%)	0148	0.6950	0.0300	0.6365	0.0522	0.0131	26	1.12	128.05	0
128	Isoleucine (%)	0675	0.7400	0.0400	0.6365	0.0522	0.0131	26	1.98	128.00	0
128	Isoleucine (%)	2188	0.7650	0.0000	0.6365	0.0522	0.0131	26	2.46	128.05	0
128	Isoleucine (%)	0226	0.5653	0.0896	0.6365	0.0522	0.0131	26	-1.36	128.00	1
129	Leucine (%)	0682	1.075	0.0000	1.243	0.0474	0.0173	27	-3.54	129.00	0
129	Leucine (%)	0652	1.150	0.0000	1.243	0.0474	0.0173	27	-1.95	129.00	0
129	Leucine (%)	0859	1.180	0.0168	1.243	0.0474	0.0173	27	-1.32	129.00	0
129	Leucine (%)	0676	1.195	0.0100	1.243	0.0474	0.0173	27	-1.01	129.05	0
129	Leucine (%)	0208	1.205	0.0100	1.243	0.0474	0.0173	27	-0.79	129.00	0
129	Leucine (%)	0918	1.206	0.0295	1.243	0.0474	0.0173	27	-0.77	129.00	0
129	Leucine (%)	2196	1.212	0.0000	1.243	0.0474	0.0173	27	-0.65	129.05	0
129	Leucine (%)	0354	1.215	0.0060	1.243	0.0474	0.0173	27	-0.58	129.00	0
129	Leucine (%)	0968	1.215	0.0300	1.243	0.0474	0.0173	27	-0.58	129.00	0
129	Leucine (%)	0626	1.218	0.0000	1.243	0.0474	0.0173	27	-0.52	129.05	0
129	Leucine (%)	0910	1.225	0.0500	1.243	0.0474	0.0173	27	-0.37	129.00	0
129	Leucine (%)	0675	1.230	0.0200	1.243	0.0474	0.0173	27	-0.27	129.00	0
129	Leucine (%)	0227	1.230	0.0200	1.243	0.0474	0.0173	27	-0.27	129.99	0
129	Leucine (%)	0870	1.240	0.0141	1.243	0.0474	0.0173	27	-0.05	129.00	0
129	Leucine (%)	0571	1.255	0.0030	1.243	0.0474	0.0173	27	0.25	129.00	0
129	Leucine (%)	0868	1.255	0.0010	1.243	0.0474	0.0173	27	0.25	129.00	0
129	Leucine (%)	0098	1.258	0.0000	1.243	0.0474	0.0173	27	0.32	129.02	0
129	Leucine (%)	0878	1.264	0.0060	1.243	0.0474	0.0173	27	0.45	129.00	0
129	Leucine (%)	2059	1.272	0.0050	1.243	0.0474	0.0173	27	0.61	129.00	0
129	Leucine (%)	0148	1.275	0.0500	1.243	0.0474	0.0173	27	0.68	129.05	0
129	Leucine (%)	0941	1.279	0.0690	1.243	0.0474	0.0173	27	0.76	129.00	0
129	Leucine (%)	0353	1.285	0.0100	1.243	0.0474	0.0173	27	0.89	129.00	0
129	Leucine (%)	0226	1.286	0.0727	1.243	0.0474	0.0173	27	0.90	129.00	0
129	Leucine (%)	0872	1.288	0.0020	1.243	0.0474	0.0173	27	0.96	129.00	0
129	Leucine (%)	0889	1.295	0.0400	1.243	0.0474	0.0173	27	1.10	129.99	0
129	Leucine (%)	2188	1.409	0.0000	1.243	0.0474	0.0173	27	3.51	129.05	0
129	Leucine (%)	0407	1.410	0.0030	1.243	0.0474	0.0173	27	3.52	129.05	0
130	L-Lysine (%)	0889	0.4775	0.0450	0.8207	0.0449	0.0210	26	-7.64	130.99	0
130	L-Lysine (%)	0682	0.7200	0.0000	0.8207	0.0449	0.0210	26	-2.24	130.00	0
130	L-Lysine (%)	0407	0.7555	0.0030	0.8207	0.0449	0.0210	26	-1.45	130.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
130	L-Lysine (%)	0859	0.7584	0.0212	0.8207	0.0449	0.0210	26	-1.39	130.00	0
130	L-Lysine (%)	0652	0.7700	0.0200	0.8207	0.0449	0.0210	26	-1.13	130.00	0
130	L-Lysine (%)	2188	0.7710	0.0000	0.8207	0.0449	0.0210	26	-1.11	130.05	0
130	L-Lysine (%)	0148	0.7950	0.0500	0.8207	0.0449	0.0210	26	-0.57	130.05	0
130	L-Lysine (%)	0208	0.7950	0.0100	0.8207	0.0449	0.0210	26	-0.57	130.00	0
130	L-Lysine (%)	0626	0.8080	0.0540	0.8207	0.0449	0.0210	26	-0.28	130.05	0
130	L-Lysine (%)	0354	0.8175	0.0030	0.8207	0.0449	0.0210	26	-0.07	130.00	0
130	L-Lysine (%)	0918	0.8200	0.0024	0.8207	0.0449	0.0210	26	-0.01	130.00	0
130	L-Lysine (%)	0968	0.8300	0.0200	0.8207	0.0449	0.0210	26	0.21	130.00	0
130	L-Lysine (%)	0353	0.8350	0.0100	0.8207	0.0449	0.0210	26	0.32	130.00	0
130	L-Lysine (%)	0868	0.8350	0.0020	0.8207	0.0449	0.0210	26	0.32	130.00	0
130	L-Lysine (%)	0872	0.8375	0.0070	0.8207	0.0449	0.0210	26	0.37	130.00	0
130	L-Lysine (%)	0878	0.8380	0.0040	0.8207	0.0449	0.0210	26	0.39	130.00	0
130	L-Lysine (%)	0941	0.8415	0.0370	0.8207	0.0449	0.0210	26	0.46	130.00	0
130	L-Lysine (%)	0910	0.8450	0.0700	0.8207	0.0449	0.0210	26	0.54	130.00	0
130	L-Lysine (%)	2059	0.8450	0.0100	0.8207	0.0449	0.0210	26	0.54	130.00	0
130	L-Lysine (%)	0571	0.8475	0.0210	0.8207	0.0449	0.0210	26	0.60	130.00	0
130	L-Lysine (%)	0676	0.8500	0.0000	0.8207	0.0449	0.0210	26	0.65	130.05	0
130	L-Lysine (%)	2196	0.8550	0.0000	0.8207	0.0449	0.0210	26	0.76	130.05	0
130	L-Lysine (%)	0675	0.8600	0.0400	0.8207	0.0449	0.0210	26	0.88	130.00	0
130	L-Lysine (%)	0098	0.8690	0.0300	0.8207	0.0449	0.0210	26	1.08	130.02	0
130	L-Lysine (%)	0870	0.8718	0.0469	0.8207	0.0449	0.0210	26	1.14	130.00	0
130	L-Lysine (%)	0227	0.8800	0.0400	0.8207	0.0449	0.0210	26	1.32	130.99	0
130	L-Lysine (%)	0226	0.8249	0.1181	0.8207	0.0449	0.0210	26	0.09	130.00	1
131	Methionine (%)	0652	0.1500	0.0200	0.2405	0.0181	0.0063	26	-5.01	131.00	0
131	Methionine (%)	0859	0.2122	0.0041	0.2405	0.0181	0.0063	26	-1.57	131.00	0
131	Methionine (%)	2188	0.2140	0.0000	0.2405	0.0181	0.0063	26	-1.47	131.05	0
131	Methionine (%)	0626	0.2195	0.0230	0.2405	0.0181	0.0063	26	-1.16	131.05	0
131	Methionine (%)	0889	0.2200	0.0100	0.2405	0.0181	0.0063	26	-1.13	131.99	0
131	Methionine (%)	0918	0.2250	0.0100	0.2405	0.0181	0.0063	26	-0.86	131.00	0
131	Methionine (%)	0941	0.2285	0.0010	0.2405	0.0181	0.0063	26	-0.66	131.00	0
131	Methionine (%)	0098	0.2345	0.0030	0.2405	0.0181	0.0063	26	-0.33	131.02	0
131	Methionine (%)	0910	0.2350	0.0100	0.2405	0.0181	0.0063	26	-0.30	131.00	0
131	Methionine (%)	0407	0.2365	0.0010	0.2405	0.0181	0.0063	26	-0.22	131.05	0
131	Methionine (%)	0227	0.2400	0.0000	0.2405	0.0181	0.0063	26	-0.03	131.99	0
131	Methionine (%)	0571	0.2415	0.0010	0.2405	0.0181	0.0063	26	0.06	131.00	0
131	Methionine (%)	0682	0.2430	0.0000	0.2405	0.0181	0.0063	26	0.14	131.00	0
131	Methionine (%)	2059	0.2430	0.0020	0.2405	0.0181	0.0063	26	0.14	131.00	0
131	Methionine (%)	0868	0.2445	0.0010	0.2405	0.0181	0.0063	26	0.22	131.00	0
131	Methionine (%)	0148	0.2450	0.0100	0.2405	0.0181	0.0063	26	0.25	131.05	0
131	Methionine (%)	0878	0.2465	0.0010	0.2405	0.0181	0.0063	26	0.33	131.00	0
131	Methionine (%)	0354	0.2475	0.0030	0.2405	0.0181	0.0063	26	0.39	131.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
131	Methionine (%)	2196	0.2480	0.0000	0.2405	0.0181	0.0063	26	0.42	131.05	0
131	Methionine (%)	0968	0.2515	0.0010	0.2405	0.0181	0.0063	26	0.61	131.00	0
131	Methionine (%)	0872	0.2535	0.0070	0.2405	0.0181	0.0063	26	0.72	131.00	0
131	Methionine (%)	0676	0.2550	0.0100	0.2405	0.0181	0.0063	26	0.81	131.05	0
131	Methionine (%)	0870	0.2580	0.0113	0.2405	0.0181	0.0063	26	0.97	131.00	0
131	Methionine (%)	2053	0.2600	0.0160	0.2405	0.0181	0.0063	26	1.08	131.99	0
131	Methionine (%)	0226	0.2900	0.0086	0.2405	0.0181	0.0063	26	2.74	131.00	0
131	Methionine (%)	0353	0.2950	0.0100	0.2405	0.0181	0.0063	26	3.02	131.00	0
131	Methionine (%)	0675	0.2850	0.0500	0.2405	0.0181	0.0063	26	2.47	131.00	1
132	Phenylalanine (%)	0652	0.5750	0.0100	0.7578	0.0428	0.0118	27	-4.28	132.00	0
132	Phenylalanine (%)	0682	0.6060	0.0000	0.7578	0.0428	0.0118	27	-3.55	132.00	0
132	Phenylalanine (%)	0889	0.6775	0.0050	0.7578	0.0428	0.0118	27	-1.88	132.99	0
132	Phenylalanine (%)	0910	0.7100	0.0400	0.7578	0.0428	0.0118	27	-1.12	132.00	0
132	Phenylalanine (%)	0675	0.7200	0.0400	0.7578	0.0428	0.0118	27	-0.88	132.00	0
132	Phenylalanine (%)	0676	0.7250	0.0100	0.7578	0.0428	0.0118	27	-0.77	132.05	0
132	Phenylalanine (%)	0226	0.7254	0.0347	0.7578	0.0428	0.0118	27	-0.76	132.00	0
132	Phenylalanine (%)	0626	0.7390	0.0000	0.7578	0.0428	0.0118	27	-0.44	132.05	0
132	Phenylalanine (%)	0354	0.7435	0.0010	0.7578	0.0428	0.0118	27	-0.33	132.00	0
132	Phenylalanine (%)	0227	0.7450	0.0100	0.7578	0.0428	0.0118	27	-0.30	132.99	0
132	Phenylalanine (%)	0870	0.7520	0.0016	0.7578	0.0428	0.0118	27	-0.14	132.00	0
132	Phenylalanine (%)	0868	0.7540	0.0020	0.7578	0.0428	0.0118	27	-0.09	132.00	0
132	Phenylalanine (%)	0968	0.7550	0.0100	0.7578	0.0428	0.0118	27	-0.07	132.00	0
132	Phenylalanine (%)	2196	0.7620	0.0000	0.7578	0.0428	0.0118	27	0.10	132.05	0
132	Phenylalanine (%)	0859	0.7641	0.0153	0.7578	0.0428	0.0118	27	0.15	132.00	0
132	Phenylalanine (%)	0571	0.7695	0.0110	0.7578	0.0428	0.0118	27	0.27	132.00	0
132	Phenylalanine (%)	0098	0.7695	0.0090	0.7578	0.0428	0.0118	27	0.27	132.02	0
132	Phenylalanine (%)	0208	0.7700	0.0200	0.7578	0.0428	0.0118	27	0.28	132.00	0
132	Phenylalanine (%)	0878	0.7705	0.0070	0.7578	0.0428	0.0118	27	0.30	132.00	0
132	Phenylalanine (%)	2059	0.7750	0.0020	0.7578	0.0428	0.0118	27	0.40	132.00	0
132	Phenylalanine (%)	0872	0.7810	0.0060	0.7578	0.0428	0.0118	27	0.54	132.00	0
132	Phenylalanine (%)	0918	0.7828	0.0279	0.7578	0.0428	0.0118	27	0.58	132.00	0
132	Phenylalanine (%)	0941	0.7960	0.0480	0.7578	0.0428	0.0118	27	0.89	132.00	0
132	Phenylalanine (%)	0148	0.8100	0.0000	0.7578	0.0428	0.0118	27	1.22	132.05	0
132	Phenylalanine (%)	2188	0.8190	0.0000	0.7578	0.0428	0.0118	27	1.43	132.05	0
132	Phenylalanine (%)	0353	0.8200	0.0000	0.7578	0.0428	0.0118	27	1.45	132.00	0
132	Phenylalanine (%)	0407	0.9030	0.0080	0.7578	0.0428	0.0118	27	3.40	132.05	0
133	Proline (%)	0652	0.8500	0.0400	0.9547	0.0526	0.0218	25	-1.99	133.00	0
133	Proline (%)	0682	0.8530	0.0000	0.9547	0.0526	0.0218	25	-1.93	133.00	0
133	Proline (%)	0354	0.8585	0.0010	0.9547	0.0526	0.0218	25	-1.83	133.00	0
133	Proline (%)	0870	0.8948	0.0173	0.9547	0.0526	0.0218	25	-1.14	133.00	0
133	Proline (%)	0676	0.9200	0.0800	0.9547	0.0526	0.0218	25	-0.66	133.05	0
133	Proline (%)	2059	0.9275	0.0010	0.9547	0.0526	0.0218	25	-0.52	133.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
133	Proline (%)	0868	0.9285	0.0050	0.9547	0.0526	0.0218	25	-0.50	133.00	0
133	Proline (%)	0889	0.9325	0.0350	0.9547	0.0526	0.0218	25	-0.42	133.99	0
133	Proline (%)	0675	0.9350	0.0500	0.9547	0.0526	0.0218	25	-0.37	133.00	0
133	Proline (%)	0208	0.9400	0.0200	0.9547	0.0526	0.0218	25	-0.28	133.00	0
133	Proline (%)	0878	0.9455	0.0010	0.9547	0.0526	0.0218	25	-0.17	133.00	0
133	Proline (%)	0571	0.9505	0.0250	0.9547	0.0526	0.0218	25	-0.08	133.00	0
133	Proline (%)	0872	0.9525	0.0010	0.9547	0.0526	0.0218	25	-0.04	133.00	0
133	Proline (%)	0626	0.9585	0.0010	0.9547	0.0526	0.0218	25	0.07	133.05	0
133	Proline (%)	0910	0.9650	0.0100	0.9547	0.0526	0.0218	25	0.20	133.00	0
133	Proline (%)	0968	0.9650	0.0100	0.9547	0.0526	0.0218	25	0.20	133.00	0
133	Proline (%)	0148	0.9750	0.0500	0.9547	0.0526	0.0218	25	0.39	133.05	0
133	Proline (%)	0407	0.9795	0.0030	0.9547	0.0526	0.0218	25	0.47	133.05	0
133	Proline (%)	2196	0.9830	0.0000	0.9547	0.0526	0.0218	25	0.54	133.05	0
133	Proline (%)	0353	0.9850	0.0100	0.9547	0.0526	0.0218	25	0.58	133.00	0
133	Proline (%)	0941	0.9910	0.0080	0.9547	0.0526	0.0218	25	0.69	133.00	0
133	Proline (%)	0227	1.010	0.0000	0.9547	0.0526	0.0218	25	1.05	133.99	0
133	Proline (%)	0918	1.082	0.1539	0.9547	0.0526	0.0218	25	2.43	133.00	0
133	Proline (%)	2188	1.182	0.0000	0.9547	0.0526	0.0218	25	4.32	133.05	0
133	Proline (%)	0859	1.260	0.0234	0.9547	0.0526	0.0218	25	5.80	133.00	0
133	Proline (%)	0226	1.056	0.1738	0.9547	0.0526	0.0218	25	1.92	133.00	1
134	Serine (%)	0652	0.6400	0.0000	0.7584	0.0411	0.0139	26	-2.88	134.00	0
134	Serine (%)	0098	0.6925	0.0410	0.7584	0.0411	0.0139	26	-1.60	134.02	0
134	Serine (%)	0676	0.6950	0.0100	0.7584	0.0411	0.0139	26	-1.54	134.05	0
134	Serine (%)	0889	0.7125	0.0050	0.7584	0.0411	0.0139	26	-1.12	134.99	0
134	Serine (%)	0682	0.7140	0.0000	0.7584	0.0411	0.0139	26	-1.08	134.00	0
134	Serine (%)	0918	0.7147	0.0701	0.7584	0.0411	0.0139	26	-1.07	134.00	0
134	Serine (%)	0859	0.7454	0.0168	0.7584	0.0411	0.0139	26	-0.32	134.00	0
134	Serine (%)	2059	0.7465	0.0130	0.7584	0.0411	0.0139	26	-0.29	134.00	0
134	Serine (%)	0208	0.7500	0.0200	0.7584	0.0411	0.0139	26	-0.20	134.00	0
134	Serine (%)	2196	0.7500	0.0000	0.7584	0.0411	0.0139	26	-0.20	134.05	0
134	Serine (%)	0571	0.7585	0.0010	0.7584	0.0411	0.0139	26	0.00	134.00	0
134	Serine (%)	0872	0.7585	0.0250	0.7584	0.0411	0.0139	26	0.00	134.00	0
134	Serine (%)	0626	0.7595	0.0050	0.7584	0.0411	0.0139	26	0.03	134.05	0
134	Serine (%)	0675	0.7600	0.0200	0.7584	0.0411	0.0139	26	0.04	134.00	0
134	Serine (%)	0227	0.7650	0.0100	0.7584	0.0411	0.0139	26	0.16	134.99	0
134	Serine (%)	0968	0.7700	0.0200	0.7584	0.0411	0.0139	26	0.28	134.00	0
134	Serine (%)	0868	0.7725	0.0030	0.7584	0.0411	0.0139	26	0.34	134.00	0
134	Serine (%)	0878	0.7745	0.0090	0.7584	0.0411	0.0139	26	0.39	134.00	0
134	Serine (%)	0941	0.7775	0.0230	0.7584	0.0411	0.0139	26	0.46	134.00	0
134	Serine (%)	0910	0.7800	0.0200	0.7584	0.0411	0.0139	26	0.53	134.00	0
134	Serine (%)	0354	0.7840	0.0020	0.7584	0.0411	0.0139	26	0.62	134.00	0
134	Serine (%)	2188	0.7850	0.0000	0.7584	0.0411	0.0139	26	0.65	134.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
134	Serine (%)	0148	0.7950	0.0300	0.7584	0.0411	0.0139	26	0.89	134.05	0
134	Serine (%)	0353	0.8150	0.0100	0.7584	0.0411	0.0139	26	1.38	134.00	0
134	Serine (%)	0226	0.8325	0.0031	0.7584	0.0411	0.0139	26	1.80	134.00	0
134	Serine (%)	0870	0.8373	0.0031	0.7584	0.0411	0.0139	26	1.92	134.00	0
134	Serine (%)	0407	0.7285	0.1570	0.7584	0.0411	0.0139	26	-0.73	134.05	1
135	Threonine (%)	0889	0.5025	0.0050	0.6026	0.0299	0.0103	26	-3.35	135.99	0
135	Threonine (%)	0652	0.5150	0.0300	0.6026	0.0299	0.0103	26	-2.93	135.00	0
135	Threonine (%)	0407	0.5180	0.0620	0.6026	0.0299	0.0103	26	-2.83	135.05	0
135	Threonine (%)	0682	0.5470	0.0000	0.6026	0.0299	0.0103	26	-1.86	135.00	0
135	Threonine (%)	0676	0.5650	0.0100	0.6026	0.0299	0.0103	26	-1.26	135.05	0
135	Threonine (%)	0626	0.5750	0.0120	0.6026	0.0299	0.0103	26	-0.92	135.05	0
135	Threonine (%)	0354	0.5880	0.0020	0.6026	0.0299	0.0103	26	-0.49	135.00	0
135	Threonine (%)	0571	0.5970	0.0020	0.6026	0.0299	0.0103	26	-0.19	135.00	0
135	Threonine (%)	2059	0.6005	0.0010	0.6026	0.0299	0.0103	26	-0.07	135.00	0
135	Threonine (%)	0910	0.6050	0.0300	0.6026	0.0299	0.0103	26	0.08	135.00	0
135	Threonine (%)	0227	0.6050	0.0100	0.6026	0.0299	0.0103	26	0.08	135.99	0
135	Threonine (%)	0872	0.6080	0.0140	0.6026	0.0299	0.0103	26	0.18	135.00	0
135	Threonine (%)	0098	0.6080	0.0060	0.6026	0.0299	0.0103	26	0.18	135.02	0
135	Threonine (%)	0868	0.6095	0.0010	0.6026	0.0299	0.0103	26	0.23	135.00	0
135	Threonine (%)	0353	0.6100	0.0000	0.6026	0.0299	0.0103	26	0.25	135.00	0
135	Threonine (%)	2196	0.6120	0.0000	0.6026	0.0299	0.0103	26	0.31	135.05	0
135	Threonine (%)	0208	0.6150	0.0100	0.6026	0.0299	0.0103	26	0.41	135.00	0
135	Threonine (%)	0968	0.6150	0.0100	0.6026	0.0299	0.0103	26	0.41	135.00	0
135	Threonine (%)	0859	0.6155	0.0159	0.6026	0.0299	0.0103	26	0.43	135.00	0
135	Threonine (%)	0878	0.6160	0.0040	0.6026	0.0299	0.0103	26	0.45	135.00	0
135	Threonine (%)	0941	0.6190	0.0240	0.6026	0.0299	0.0103	26	0.55	135.00	0
135	Threonine (%)	0870	0.6233	0.0051	0.6026	0.0299	0.0103	26	0.69	135.00	0
135	Threonine (%)	0226	0.6260	0.0039	0.6026	0.0299	0.0103	26	0.78	135.00	0
135	Threonine (%)	0148	0.6300	0.0000	0.6026	0.0299	0.0103	26	0.91	135.05	0
135	Threonine (%)	2188	0.6510	0.0000	0.6026	0.0299	0.0103	26	1.62	135.05	0
135	Threonine (%)	0675	0.6750	0.0100	0.6026	0.0299	0.0103	26	2.42	135.00	0
135	Threonine (%)	0918	0.6334	0.0965	0.6026	0.0299	0.0103	26	1.03	135.00	1
136	Tryptophan (%)	0208	0.1635	0.0090	0.2019	0.0249	0.0050	16	-1.54	136.01	0
136	Tryptophan (%)	0910	0.1750	0.0100	0.2019	0.0249	0.0050	16	-1.08	136.00	0
136	Tryptophan (%)	0941	0.1785	0.0090	0.2019	0.0249	0.0050	16	-0.94	136.00	0
136	Tryptophan (%)	0870	0.1841	0.0070	0.2019	0.0249	0.0050	16	-0.71	136.00	0
136	Tryptophan (%)	0682	0.1940	0.0000	0.2019	0.0249	0.0050	16	-0.32	136.00	0
136	Tryptophan (%)	0098	0.1960	0.0000	0.2019	0.0249	0.0050	16	-0.24	136.02	0
136	Tryptophan (%)	0872	0.1975	0.0010	0.2019	0.0249	0.0050	16	-0.18	136.03	0
136	Tryptophan (%)	0859	0.1979	0.0064	0.2019	0.0249	0.0050	16	-0.16	136.03	0
136	Tryptophan (%)	0878	0.1980	0.0000	0.2019	0.0249	0.0050	16	-0.16	136.01	0
136	Tryptophan (%)	2059	0.2040	0.0020	0.2019	0.0249	0.0050	16	0.09	136.03	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
136	Tryptophan (%)	0571	0.2050	0.0020	0.2019	0.0249	0.0050	16	0.13	136.03	0
136	Tryptophan (%)	0868	0.2070	0.0080	0.2019	0.0249	0.0050	16	0.21	136.01	0
136	Tryptophan (%)	0353	0.2150	0.0100	0.2019	0.0249	0.0050	16	0.53	136.03	0
136	Tryptophan (%)	0918	0.2350	0.0100	0.2019	0.0249	0.0050	16	1.33	136.00	0
136	Tryptophan (%)	0227	0.2400	0.0000	0.2019	0.0249	0.0050	16	1.53	136.00	0
136	Tryptophan (%)	0889	0.4425	0.0050	0.2019	0.0249	0.0050	16	9.67	136.99	0
136	Tryptophan (%)	0407	0.2675	0.0390	0.2019	0.0249	0.0050	16	2.64	136.05	1
137	Tyrosine (%)	0910	0.3750	0.0100	0.5203	0.0801	0.0177	21	-1.81	137.00	0
137	Tyrosine (%)	0682	0.4200	0.0000	0.5203	0.0801	0.0177	21	-1.25	137.00	0
137	Tyrosine (%)	0870	0.4277	0.0126	0.5203	0.0801	0.0177	21	-1.16	137.00	0
137	Tyrosine (%)	0889	0.4425	0.0050	0.5203	0.0801	0.0177	21	-0.97	137.99	0
137	Tyrosine (%)	0227	0.4550	0.0500	0.5203	0.0801	0.0177	21	-0.82	137.99	0
137	Tyrosine (%)	0626	0.4590	0.0620	0.5203	0.0801	0.0177	21	-0.77	137.05	0
137	Tyrosine (%)	0354	0.4820	0.0040	0.5203	0.0801	0.0177	21	-0.48	137.00	0
137	Tyrosine (%)	0968	0.4850	0.0300	0.5203	0.0801	0.0177	21	-0.44	137.00	0
137	Tyrosine (%)	0098	0.5065	0.0010	0.5203	0.0801	0.0177	21	-0.17	137.02	0
137	Tyrosine (%)	0918	0.5170	0.0032	0.5203	0.0801	0.0177	21	-0.04	137.00	0
137	Tyrosine (%)	0676	0.5200	0.0200	0.5203	0.0801	0.0177	21	0.00	137.05	0
137	Tyrosine (%)	0226	0.5401	0.0127	0.5203	0.0801	0.0177	21	0.25	137.00	0
137	Tyrosine (%)	0941	0.5445	0.0650	0.5203	0.0801	0.0177	21	0.30	137.00	0
137	Tyrosine (%)	0872	0.5490	0.0060	0.5203	0.0801	0.0177	21	0.36	137.00	0
137	Tyrosine (%)	0675	0.5500	0.0000	0.5203	0.0801	0.0177	21	0.37	137.00	0
137	Tyrosine (%)	2196	0.5710	0.0000	0.5203	0.0801	0.0177	21	0.63	137.05	0
137	Tyrosine (%)	2059	0.5805	0.0010	0.5203	0.0801	0.0177	21	0.75	137.00	0
137	Tyrosine (%)	0353	0.5950	0.0100	0.5203	0.0801	0.0177	21	0.93	137.00	0
137	Tyrosine (%)	0208	0.6000	0.0000	0.5203	0.0801	0.0177	21	1.00	137.00	0
137	Tyrosine (%)	0148	0.6600	0.0800	0.5203	0.0801	0.0177	21	1.75	137.05	0
137	Tyrosine (%)	2188	0.6810	0.0000	0.5203	0.0801	0.0177	21	2.01	137.05	0
137	Tyrosine (%)	0407	1.947	0.7250	0.5203	0.0801	0.0177	21	17.81	137.05	2
138	Valine (%)	0889	0.6700	0.0000	0.7524	0.0506	0.0138	26	-1.63	138.99	0
138	Valine (%)	0682	0.6720	0.0000	0.7524	0.0506	0.0138	26	-1.59	138.00	0
138	Valine (%)	0859	0.6842	0.0143	0.7524	0.0506	0.0138	26	-1.35	138.00	0
138	Valine (%)	0918	0.7113	0.0277	0.7524	0.0506	0.0138	26	-0.81	138.00	0
138	Valine (%)	0868	0.7135	0.0070	0.7524	0.0506	0.0138	26	-0.77	138.00	0
138	Valine (%)	0354	0.7165	0.0070	0.7524	0.0506	0.0138	26	-0.71	138.00	0
138	Valine (%)	2196	0.7250	0.0000	0.7524	0.0506	0.0138	26	-0.54	138.05	0
138	Valine (%)	0626	0.7265	0.0070	0.7524	0.0506	0.0138	26	-0.51	138.05	0
138	Valine (%)	0353	0.7300	0.0000	0.7524	0.0506	0.0138	26	-0.44	138.00	0
138	Valine (%)	0870	0.7360	0.0014	0.7524	0.0506	0.0138	26	-0.32	138.00	0
138	Valine (%)	0941	0.7375	0.0570	0.7524	0.0506	0.0138	26	-0.29	138.00	0
138	Valine (%)	0652	0.7400	0.0000	0.7524	0.0506	0.0138	26	-0.25	138.00	0
138	Valine (%)	0910	0.7400	0.0200	0.7524	0.0506	0.0138	26	-0.25	138.00	0



Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
138	Valine (%)	0676	0.7500	0.0200	0.7524	0.0506	0.0138	26	-0.05	138.05	0
138	Valine (%)	0227	0.7550	0.0300	0.7524	0.0506	0.0138	26	0.05	138.99	0
138	Valine (%)	0208	0.7600	0.0000	0.7524	0.0506	0.0138	26	0.15	138.00	0
138	Valine (%)	0878	0.7660	0.0140	0.7524	0.0506	0.0138	26	0.27	138.00	0
138	Valine (%)	0098	0.7730	0.0100	0.7524	0.0506	0.0138	26	0.41	138.02	0
138	Valine (%)	0571	0.7775	0.0010	0.7524	0.0506	0.0138	26	0.50	138.00	0
138	Valine (%)	2059	0.7885	0.0110	0.7524	0.0506	0.0138	26	0.71	138.00	0
138	Valine (%)	0968	0.7900	0.0000	0.7524	0.0506	0.0138	26	0.74	138.00	0
138	Valine (%)	0872	0.7945	0.0330	0.7524	0.0506	0.0138	26	0.83	138.00	0
138	Valine (%)	0148	0.8100	0.0600	0.7524	0.0506	0.0138	26	1.14	138.05	0
138	Valine (%)	2188	0.8290	0.0000	0.7524	0.0506	0.0138	26	1.51	138.05	0
138	Valine (%)	0675	0.8350	0.0100	0.7524	0.0506	0.0138	26	1.63	138.00	0
138	Valine (%)	0407	0.8380	0.0280	0.7524	0.0506	0.0138	26	1.69	138.05	0
138	Valine (%)	0226	0.5653	0.0896	0.7524	0.0506	0.0138	26	-3.70	138.00	1
139	Taurine (%)	0407	0.0035	0.0010				2		139.05	0
139	Taurine (%)	0682	0.0840	0.0000				2		139.00	0
139	Taurine (%)	0098	< 0.01					2		139.02	5
139	Taurine (%)	0227	< 0.01					2		139.99	5
160	Fructose (%)	0297	0.4205	0.0310				2		160.10	0
160	Fructose (%)	0227	0.6800	0.0600				2		160.99	0
161	Galactose (%)	0297	0.0000	0.0000				0		161.10	4
162	Glucose (%)	0297	0.2910	0.0060				2		162.10	0
162	Glucose (%)	0227	0.3850	0.0100				2		162.99	0
163	Lactose (%)	0297	0.4405	0.0330				2		163.10	0
163	Lactose (%)	0227	0.5200	0.0000				2		163.99	0
164	Maltose (%)	0297	0.0565	0.0050				1		164.10	0
164	Maltose (%)	0227	< 0.15					1		164.99	5
165	Sucrose (%)	0297	3.406	0.0260				2		165.10	0
165	Sucrose (%)	0227	3.625	0.0100				2		165.99	0
166	Raffinose (%)	0227	0.2750	0.0100				2		166.99	0
166	Raffinose (%)	0297	0.3040	0.0900				2		166.10	0
167	Stachyose (%)	0227	0.9850	0.0100				2		167.99	0
167	Stachyose (%)	0297	1.122	0.1940				2		167.10	0
350	Carbadox (ppm)	2055	0.5550	0.1700				2		350.03	0
350	Carbadox (ppm)	0038	0.6508	0.1315				2		350.03	0
351	Chlortetracycline (ppm)	2314	0.2268	0.0072	2.261	1.733	0.1093	4		351.05	0
351	Chlortetracycline (ppm)	2055	2.150	0.1000	2.261	1.733	0.1093	4		351.05	0
351	Chlortetracycline (ppm)	0941	2.200	0.0000	2.261	1.733	0.1093	4		351.05	0
351	Chlortetracycline (ppm)	0148	4.465	0.3300	2.261	1.733	0.1093	4		351.05	0
354	Decoquinatate (ppm)	0941	38.90	0.0000	48.80	4.927	1.374	13	-2.01	354.04	0
354	Decoquinatate (ppm)	0590	42.55	1.900	48.80	4.927	1.374	13	-1.27	354.01	0
354	Decoquinatate (ppm)	0148	44.81	0.2880	48.80	4.927	1.374	13	-0.81	354.04	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
354	Decoquinatate (ppm)	0038	46.12	1.549	48.80	4.927	1.374	13	-0.54	354.01	0
354	Decoquinatate (ppm)	0028	46.95	4.500	48.80	4.927	1.374	13	-0.38	354.01	0
354	Decoquinatate (ppm)	0036	49.00	0.6000	48.80	4.927	1.374	13	0.04	354.02	0
354	Decoquinatate (ppm)	0001	49.45	2.280	48.80	4.927	1.374	13	0.13	354.02	0
354	Decoquinatate (ppm)	0010	49.50	3.400	48.80	4.927	1.374	13	0.14	354.01	0
354	Decoquinatate (ppm)	2192	50.05	0.0900	48.80	4.927	1.374	13	0.25	354.01	0
354	Decoquinatate (ppm)	0014	51.95	0.9000	48.80	4.927	1.374	13	0.64	354.01	0
354	Decoquinatate (ppm)	0910	52.95	0.1000	48.80	4.927	1.374	13	0.84	354.01	0
354	Decoquinatate (ppm)	2113	53.50	1.000	48.80	4.927	1.374	13	0.95	354.04	0
354	Decoquinatate (ppm)	0009	76.46	1.260	48.80	4.927	1.374	13	5.61	354.01	0
361	Lasalocid Sodium (ppm)	2113	0.3950	0.1900				1		361.05	0
365	Monensin (ppm)	0009	17.75	0.3600	20.59	2.156	0.5511	16	-1.32	365.05	0
365	Monensin (ppm)	0001	17.80	2.000	20.59	2.156	0.5511	16	-1.29	365.05	0
365	Monensin (ppm)	0910	18.55	0.5000	20.59	2.156	0.5511	16	-0.95	365.99	0
365	Monensin (ppm)	0941	18.64	0.0000	20.59	2.156	0.5511	16	-0.91	365.05	0
365	Monensin (ppm)	0010	19.20	1.000	20.59	2.156	0.5511	16	-0.65	365.03	0
365	Monensin (ppm)	0036	19.55	1.700	20.59	2.156	0.5511	16	-0.48	365.03	0
365	Monensin (ppm)	0035	20.04	0.2000	20.59	2.156	0.5511	16	-0.26	365.03	0
365	Monensin (ppm)	2053	20.45	0.3000	20.59	2.156	0.5511	16	-0.07	365.04	0
365	Monensin (ppm)	0148	21.01	0.4200	20.59	2.156	0.5511	16	0.19	365.03	0
365	Monensin (ppm)	0668	21.40	0.0000	20.59	2.156	0.5511	16	0.37	365.02	0
365	Monensin (ppm)	0038	21.46	0.6839	20.59	2.156	0.5511	16	0.40	365.05	0
365	Monensin (ppm)	2113	21.50	1.000	20.59	2.156	0.5511	16	0.42	365.05	0
365	Monensin (ppm)	0512	22.70	0.0800	20.59	2.156	0.5511	16	0.98	365.99	0
365	Monensin (ppm)	2192	22.74	0.0100	20.59	2.156	0.5511	16	0.99	365.02	0
365	Monensin (ppm)	0148	22.86	0.2640	20.59	2.156	0.5511	16	1.05	365.05	0
365	Monensin (ppm)	0884	25.25	0.3000	20.59	2.156	0.5511	16	2.16	365.05	0
365	Monensin (ppm)	0590	21.66	3.880	20.59	2.156	0.5511	16	0.50	365.03	1
382	Sulfamethazine (ppm)	2314	0.1416	0.0070				3		382.04	0
382	Sulfamethazine (ppm)	0941	0.3430	0.0000				3		382.04	0
382	Sulfamethazine (ppm)	2055	0.4000	0.0000				3		382.04	0
382	Sulfamethazine (ppm)	0148	0.2725	0.0390				3		382.04	1
386	Tiamulin (ppm)	2314	0.0864	0.0064	0.1502	0.0787	0.0364	4		386.02	0
386	Tiamulin (ppm)	0941	0.1220	0.0000	0.1502	0.0787	0.0364	4		386.02	0
386	Tiamulin (ppm)	0148	0.1275	0.0890	0.1502	0.0787	0.0364	4		386.02	0
386	Tiamulin (ppm)	2055	0.2650	0.0500	0.1502	0.0787	0.0364	4		386.02	0
400	Water Activity (Units)	2174	0.5360	0.0120	0.5542	0.0181	0.0067	13	-1.00	400.01	0
400	Water Activity (Units)	0942	0.5380	0.0060	0.5542	0.0181	0.0067	13	-0.89	400.01	0
400	Water Activity (Units)	2302	0.5385	0.0150	0.5542	0.0181	0.0067	13	-0.87	400.99	0
400	Water Activity (Units)	0843	0.5415	0.0030	0.5542	0.0181	0.0067	13	-0.70	400.99	0
400	Water Activity (Units)	0407	0.5457	0.0068	0.5542	0.0181	0.0067	13	-0.47	400.01	0
400	Water Activity (Units)	2033	0.5460	0.0020	0.5542	0.0181	0.0067	13	-0.45	400.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
400	Water Activity (Units)	2168	0.5475	0.0250	0.5542	0.0181	0.0067	13	-0.37	400.01	0
400	Water Activity (Units)	0843	0.5557	0.0016	0.5542	0.0181	0.0067	13	0.09	400.01	0
400	Water Activity (Units)	0589	0.5580	0.0040	0.5542	0.0181	0.0067	13	0.21	400.01	0
400	Water Activity (Units)	2303	0.5617	0.0009	0.5542	0.0181	0.0067	13	0.41	400.01	0
400	Water Activity (Units)	2181	0.5752	0.0024	0.5542	0.0181	0.0067	13	1.16	400.01	0
400	Water Activity (Units)	2073	0.5790	0.0020	0.5542	0.0181	0.0067	13	1.38	400.99	0
400	Water Activity (Units)	2109	0.5989	0.0068	0.5542	0.0181	0.0067	13	2.48	400.01	0
516	Arsenic, Total (ppm)	0045	0.2255	0.0050	0.2791	0.0236	0.0251	13	-2.27	516.00	0
516	Arsenic, Total (ppm)	0171	0.2575	0.0070	0.2791	0.0236	0.0251	13	-0.91	516.00	0
516	Arsenic, Total (ppm)	0910	0.2650	0.0300	0.2791	0.0236	0.0251	13	-0.60	516.52	0
516	Arsenic, Total (ppm)	0186	0.2665	0.0030	0.2791	0.0236	0.0251	13	-0.53	516.52	0
516	Arsenic, Total (ppm)	0096	0.2675	0.0250	0.2791	0.0236	0.0251	13	-0.49	516.52	0
516	Arsenic, Total (ppm)	0199	0.2700	0.0800	0.2791	0.0236	0.0251	13	-0.38	516.53	0
516	Arsenic, Total (ppm)	0553	0.2770	0.0020	0.2791	0.0236	0.0251	13	-0.09	516.53	0
516	Arsenic, Total (ppm)	0227	0.2820	0.0120	0.2791	0.0236	0.0251	13	0.12	516.53	0
516	Arsenic, Total (ppm)	0939	0.2850	0.0100	0.2791	0.0236	0.0251	13	0.25	516.53	0
516	Arsenic, Total (ppm)	0941	0.2920	0.0000	0.2791	0.0236	0.0251	13	0.55	516.43	0
516	Arsenic, Total (ppm)	2012	0.2930	0.0040	0.2791	0.0236	0.0251	13	0.59	516.53	0
516	Arsenic, Total (ppm)	0918	0.3144	0.0069	0.2791	0.0236	0.0251	13	1.49	516.53	0
516	Arsenic, Total (ppm)	0407	1.436	0.1420	0.2791	0.0236	0.0251	13	49.03	516.43	0
516	Arsenic, Total (ppm)	2302	0.0500	0.1000	0.2791	0.0236	0.0251	13		516.99	4
516	Arsenic, Total (ppm)	0682	< 10		0.2791	0.0236	0.0251	13		516.43	5
518	Cadmium (ppm)	0171	0.0800	0.0020	0.1093	0.0103	0.0085	14	-2.85	518.41	0
518	Cadmium (ppm)	0407	0.0945	0.0012	0.1093	0.0103	0.0085	14	-1.44	518.41	0
518	Cadmium (ppm)	0199	0.1000	0.0200	0.1093	0.0103	0.0085	14	-0.91	518.53	0
518	Cadmium (ppm)	0939	0.1050	0.0300	0.1093	0.0103	0.0085	14	-0.42	518.53	0
518	Cadmium (ppm)	0910	0.1050	0.0100	0.1093	0.0103	0.0085	14	-0.42	518.52	0
518	Cadmium (ppm)	0186	0.1065	0.0010	0.1093	0.0103	0.0085	14	-0.27	518.52	0
518	Cadmium (ppm)	0096	0.1070	0.0020	0.1093	0.0103	0.0085	14	-0.23	518.52	0
518	Cadmium (ppm)	2012	0.1125	0.0010	0.1093	0.0103	0.0085	14	0.31	518.53	0
518	Cadmium (ppm)	0918	0.1157	0.0068	0.1093	0.0103	0.0085	14	0.62	518.53	0
518	Cadmium (ppm)	0227	0.1160	0.0060	0.1093	0.0103	0.0085	14	0.65	518.53	0
518	Cadmium (ppm)	0941	0.1160	0.0000	0.1093	0.0103	0.0085	14	0.65	518.43	0
518	Cadmium (ppm)	0968	0.1160	0.0020	0.1093	0.0103	0.0085	14	0.65	518.43	0
518	Cadmium (ppm)	0553	0.1195	0.0010	0.1093	0.0103	0.0085	14	0.99	518.53	0
518	Cadmium (ppm)	0407	0.1231	0.0359	0.1093	0.0103	0.0085	14	1.33	518.43	0
518	Cadmium (ppm)	2302	17.00	0.0000	0.1093	0.0103	0.0085	14	1640.01	518.99	2
518	Cadmium (ppm)	0682	< 0.2		0.1093	0.0103	0.0085	14		518.43	5
520	Chromium (ppm)	0171	2.165	0.0100	4.510	1.745	0.2391	14	-1.34	520.41	0
520	Chromium (ppm)	0968	2.296	0.0200	4.510	1.745	0.2391	14	-1.27	520.43	0
520	Chromium (ppm)	0186	2.519	0.0410	4.510	1.745	0.2391	14	-1.14	520.52	0
520	Chromium (ppm)	0407	2.928	0.1780	4.510	1.745	0.2391	14	-0.91	520.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
520	Chromium (ppm)	0918	3.836	0.4069	4.510	1.745	0.2391	14	-0.39	520.53	0
520	Chromium (ppm)	0096	3.941	0.1760	4.510	1.745	0.2391	14	-0.33	520.52	0
520	Chromium (ppm)	0096	4.743	0.4940	4.510	1.745	0.2391	14	0.13	520.42	0
520	Chromium (ppm)	0939	5.345	0.2500	4.510	1.745	0.2391	14	0.48	520.53	0
520	Chromium (ppm)	0164	5.385	0.1500	4.510	1.745	0.2391	14	0.50	520.53	0
520	Chromium (ppm)	0553	5.615	0.0300	4.510	1.745	0.2391	14	0.63	520.53	0
520	Chromium (ppm)	0682	5.740	0.0000	4.510	1.745	0.2391	14	0.71	520.43	0
520	Chromium (ppm)	0693	5.807	1.221	4.510	1.745	0.2391	14	0.74	520.42	0
520	Chromium (ppm)	0510	6.385	0.0300	4.510	1.745	0.2391	14	1.07	520.43	0
520	Chromium (ppm)	0045	6.430	0.3400	4.510	1.745	0.2391	14	1.10	520.42	0
520	Chromium (ppm)	0407	4.677	1.718	4.510	1.745	0.2391	14	0.10	520.43	1
526	Lead (ppm)	0171	0.1260	0.0240	0.1788	0.0354	0.0230	12	-1.49	526.41	0
526	Lead (ppm)	2012	0.1465	0.0030	0.1788	0.0354	0.0230	12	-0.91	526.53	0
526	Lead (ppm)	0199	0.1500	0.0000	0.1788	0.0354	0.0230	12	-0.81	526.53	0
526	Lead (ppm)	0407	0.1597	0.0663	0.1788	0.0354	0.0230	12	-0.54	526.41	0
526	Lead (ppm)	0910	0.1650	0.0100	0.1788	0.0354	0.0230	12	-0.39	526.52	0
526	Lead (ppm)	0096	0.1750	0.0160	0.1788	0.0354	0.0230	12	-0.11	526.52	0
526	Lead (ppm)	0227	0.1860	0.0060	0.1788	0.0354	0.0230	12	0.20	526.53	0
526	Lead (ppm)	0941	0.1900	0.0000	0.1788	0.0354	0.0230	12	0.31	526.43	0
526	Lead (ppm)	0553	0.1945	0.0050	0.1788	0.0354	0.0230	12	0.44	526.53	0
526	Lead (ppm)	0918	0.2003	0.0042	0.1788	0.0354	0.0230	12	0.61	526.53	0
526	Lead (ppm)	0407	0.2212	0.0510	0.1788	0.0354	0.0230	12	1.19	526.43	0
526	Lead (ppm)	0939	0.2450	0.0900	0.1788	0.0354	0.0230	12	1.87	526.53	0
526	Lead (ppm)	2302	2.550	4.900	0.1788	0.0354	0.0230	12	66.89	526.99	1
526	Lead (ppm)	0682	< 3		0.1788	0.0354	0.0230	12		526.43	5
529	Mercury (ppb)	2012	1.618	0.4230				3		529.99	0
529	Mercury (ppb)	0941	2.000	0.0000				3		529.99	0
529	Mercury (ppb)	0199	3.000	0.0000				3		529.99	0
529	Mercury (ppb)	2302	0.0500	0.1000				3		529.99	4
529	Mercury (ppb)	0227	< 10					3		529.99	5
529	Mercury (ppb)	0682	< 10000					3		529.99	5
539	Nickel (ppm)	0186	2.109	0.0340	2.784	0.5581	0.0899	9	-1.21	539.52	0
539	Nickel (ppm)	0171	2.245	0.0300	2.784	0.5581	0.0899	9	-0.97	539.41	0
539	Nickel (ppm)	0407	2.446	0.2851	2.784	0.5581	0.0899	9	-0.61	539.41	0
539	Nickel (ppm)	0407	2.655	0.1751	2.784	0.5581	0.0899	9	-0.23	539.43	0
539	Nickel (ppm)	0941	2.750	0.0000	2.784	0.5581	0.0899	9	-0.06	539.43	0
539	Nickel (ppm)	0096	2.886	0.1110	2.784	0.5581	0.0899	9	0.18	539.52	0
539	Nickel (ppm)	0918	3.012	0.1340	2.784	0.5581	0.0899	9	0.41	539.53	0
539	Nickel (ppm)	0553	3.330	0.0400	2.784	0.5581	0.0899	9	0.98	539.53	0
539	Nickel (ppm)	0682	3.620	0.0000	2.784	0.5581	0.0899	9	1.50	539.43	0
710	Lauric Acid (12:0) (% (w/w))	0676	0.0015	0.0010				1		710.99	0
710	Lauric Acid (12:0) (% (w/w))	0226	< 0.005					1		710.99	5

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
710	Lauric Acid (12:0) (% (w/w))	0164	< 0.02						1	710.99	5
714	Myristic Acid (14:0) (% (w/w))	0226	0.0085	0.0010					2	714.99	0
714	Myristic Acid (14:0) (% (w/w))	0297	0.0179	0.0000					2	714.02	0
714	Myristic Acid (14:0) (% (w/w))	0164	< 0.02						2	714.99	5
716	Palmitic Acid (16:0) (% (w/w))	0297	0.6970	0.0000					3	716.02	0
716	Palmitic Acid (16:0) (% (w/w))	0226	0.7300	0.0220					3	716.99	0
716	Palmitic Acid (16:0) (% (w/w))	0164	0.7500	0.0200					3	716.99	0
718	Palmitoleic Acid (9c-16:1) (% (w/w))	0297	0.0090	0.0000					3	718.02	0
718	Palmitoleic Acid (9c-16:1) (% (w/w))	0676	0.0100	0.0000					3	718.99	0
718	Palmitoleic Acid (9c-16:1) (% (w/w))	0226	0.0115	0.0010					3	718.99	0
720	Margaric acid (17:0) (% (w/w))	0297	0.0080	0.0000					1	720.02	0
722	Stearic Acid (18:0) (% (w/w))	0164	0.1650	0.0100					3	722.99	0
722	Stearic Acid (18:0) (% (w/w))	0226	0.1700	0.0060					3	722.99	0
722	Stearic Acid (18:0) (% (w/w))	0297	0.1714	0.0000					3	722.02	0
724	Oleic Acid (9c-18:1) (% (w/w))	0164	1.025	0.0100					3	724.99	0
724	Oleic Acid (9c-18:1) (% (w/w))	0226	1.057	0.0040					3	724.99	0
724	Oleic Acid (9c-18:1) (% (w/w))	0297	1.101	0.0000					3	724.02	0
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0164	2.260	0.0000	2.441	0.2424	0.0223		4	726.99	0
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0226	2.324	0.0220	2.441	0.2424	0.0223		4	726.99	0
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0297	2.382	0.0000	2.441	0.2424	0.0223		4	726.02	0
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0676	2.797	0.0670	2.441	0.2424	0.0223		4	726.99	0
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v))	0226	0.2850	0.0000					3	728.99	0
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v))	0164	0.2900	0.0000					3	728.99	0
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v))	0297	0.2960	0.0000					3	728.02	0
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v))	0676	0.3200	0.0080					3	728.99	1
730	Arachidic Acid (20:0) (% (w/w))	0226	0.0160	0.0000					2	730.99	0
730	Arachidic Acid (20:0) (% (w/w))	0164	0.0200	0.0000					2	730.99	0
730	Arachidic Acid (20:0) (% (w/w))	0297	0.0000	0.0000					2	730.02	4
732	Gondoic Acid (11c-20:1) (% (w/w))	0226	0.0160	0.0020					1	732.99	0
732	Gondoic Acid (11c-20:1) (% (w/w))	0164	< 0.02						1	732.99	5
736	Arachidonic Acid (5c,8c,11c,14c-20:4) (% (w/v))	0676	0.0000	0.0000					0	736.99	4
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c) (% (w/v))	0226	0.0000	0.0000					0	740.99	4
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c) (% (w/v))	0676	0.0000	0.0000					0	740.99	4
742	Behenic Acid (22:0) (% (w/w))	0226	0.0165	0.0010					1	742.99	0
742	Behenic Acid (22:0) (% (w/w))	0164	< 0.02						1	742.99	5
744	Erucic Acid (13c-22:1) (% (w/w))	0297	0.0000	0.0000					0	744.02	4
744	Erucic Acid (13c-22:1) (% (w/w))	0226	< 0.005						0	744.99	5
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c) (% (w/v))	0226	0.0000	0.0000					0	746.99	4
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c) (% (w/v))	0676	0.0000	0.0000					0	746.99	4
748	Lignoceric Acid (24:0) (% (w/w))	0226	0.0110	0.0020					1	748.99	0
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22c) (% (w/v))	0676	0.0000	0.0000					0	750.99	4
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22c) (% (w/v))	0226	< 0.005						0	750.99	5

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
752	Nervonic Acid (24:1) isomers (% (w/w))	0226	< 0.005						0	752.99	5
754	Total n-3 Polyunsaturated (Omega-3) Fatty Aci	0164	0.2900	0.0000					2	754.99	0
754	Total n-3 Polyunsaturated (Omega-3) Fatty Aci	0676	0.3200	0.0000					2	754.99	0
756	Total n-6 Polyunsaturated (Omega-6) Fatty Aci	0164	2.270	0.0000					2	756.99	0
756	Total n-6 Polyunsaturated (Omega-6) Fatty Aci	0676	2.795	0.0700					2	756.99	0
758	Total Saturated Fatty Acids (% (w/w))	0164	0.9900	0.0200					2	758.99	0
758	Total Saturated Fatty Acids (% (w/w))	0297	4.783	0.0000					2	758.02	0
762	Total Monounsaturated Fatty Acids (% (w/w))	0297	0.8900	0.0000					2	762.02	0
762	Total Monounsaturated Fatty Acids (% (w/w))	0164	1.125	0.0100					2	762.99	0
766	Total Polyunsaturated Fatty Acids (% (w/w))	0297	1.125	0.0000					2	766.02	0
766	Total Polyunsaturated Fatty Acids (% (w/w))	0164	2.565	0.0100					2	766.99	0
770	Total Fat (equivalent to NLEA) (% (w/w))	0164	4.915	0.0500					1	770.99	0
772	Total Fatty Acids (% (w/w))	0297	2.764	0.0000					3	772.02	0
772	Total Fatty Acids (% (w/w))	0226	4.646	0.0570					3	772.99	0
772	Total Fatty Acids (% (w/w))	0164	4.700	0.0400					3	772.99	0

Notes: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2 and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = Rejected for duplicates too far apart, 2 = Rejected as outlier, 5 = A reporting limit and 4 = zeros submitted as values. Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). Flag 3 indicates not used in statistics.