

Animal Feed Scheme

Pig Grower, Medicated

Test Material Code # 202026

Labs Reporting: 177

Analytes Reported 117

Issue Date : 07/31/2020

Analyte Proficiency Testing Report

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
000	Urea (%)	0278	0.3000	0.0000				1		000.02	0
001	Loss on Drying (%)	0618	7.030	0.1000	9.426	0.3403	0.1178	76	-7.04	001.99	0
001	Loss on Drying (%)	0345	7.405	0.0700	9.426	0.3403	0.1178	76	-5.94	001.07	0
001	Loss on Drying (%)	2155	8.510	0.4200	9.426	0.3403	0.1178	76	-2.69	001.99	0
001	Loss on Drying (%)	0038	8.575	0.1100	9.426	0.3403	0.1178	76	-2.50	001.07	0
001	Loss on Drying (%)	0889	8.595	0.0700	9.426	0.3403	0.1178	76	-2.44	001.99	0
001	Loss on Drying (%)	0019	8.645	0.3100	9.426	0.3403	0.1178	76	-2.30	001.99	0
001	Loss on Drying (%)	0918	8.885	0.4100	9.426	0.3403	0.1178	76	-1.59	001.99	0
001	Loss on Drying (%)	0852	8.945	0.0700	9.426	0.3403	0.1178	76	-1.41	001.99	0
001	Loss on Drying (%)	0034	8.960	0.1400	9.426	0.3403	0.1178	76	-1.37	001.07	0
001	Loss on Drying (%)	0510	9.000	0.0000	9.426	0.3403	0.1178	76	-1.25	001.99	0
001	Loss on Drying (%)	0676	9.075	0.1900	9.426	0.3403	0.1178	76	-1.03	001.99	0
001	Loss on Drying (%)	2268	9.080	0.0800	9.426	0.3403	0.1178	76	-1.02	001.99	0
001	Loss on Drying (%)	0353	9.110	0.0800	9.426	0.3403	0.1178	76	-0.93	001.07	0
001	Loss on Drying (%)	2146	9.115	0.0900	9.426	0.3403	0.1178	76	-0.92	001.07	0
001	Loss on Drying (%)	0169	9.155	0.0100	9.426	0.3403	0.1178	76	-0.80	001.00	0
001	Loss on Drying (%)	0948	9.165	0.0500	9.426	0.3403	0.1178	76	-0.77	001.99	0
001	Loss on Drying (%)	2144	9.170	0.0200	9.426	0.3403	0.1178	76	-0.75	001.07	0
001	Loss on Drying (%)	0590	9.175	0.2300	9.426	0.3403	0.1178	76	-0.74	001.07	0
001	Loss on Drying (%)	2081	9.175	0.5700	9.426	0.3403	0.1178	76	-0.74	001.99	0
001	Loss on Drying (%)	0142	9.225	0.0100	9.426	0.3403	0.1178	76	-0.59	001.07	0
001	Loss on Drying (%)	0675	9.230	0.0200	9.426	0.3403	0.1178	76	-0.58	001.07	0
001	Loss on Drying (%)	2246	9.255	0.1300	9.426	0.3403	0.1178	76	-0.50	001.99	0
001	Loss on Drying (%)	2295	9.265	0.0100	9.426	0.3403	0.1178	76	-0.47	001.03	0
001	Loss on Drying (%)	2012	9.285	0.0300	9.426	0.3403	0.1178	76	-0.42	001.07	0
001	Loss on Drying (%)	0662	9.297	0.0066	9.426	0.3403	0.1178	76	-0.38	001.99	0
001	Loss on Drying (%)	0876	9.300	0.2000	9.426	0.3403	0.1178	76	-0.37	001.07	0
001	Loss on Drying (%)	0878	9.320	0.1200	9.426	0.3403	0.1178	76	-0.31	001.03	0
001	Loss on Drying (%)	0148	9.320	0.4400	9.426	0.3403	0.1178	76	-0.31	001.07	0
001	Loss on Drying (%)	0536	9.320	0.0200	9.426	0.3403	0.1178	76	-0.31	001.99	0
001	Loss on Drying (%)	0571	9.325	0.3900	9.426	0.3403	0.1178	76	-0.30	001.07	0
001	Loss on Drying (%)	0683	9.330	0.0200	9.426	0.3403	0.1178	76	-0.28	001.07	0
001	Loss on Drying (%)	0027	9.339	0.0240	9.426	0.3403	0.1178	76	-0.26	001.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
001	Loss on Drying (%)	0083	9.360	0.0800	9.426	0.3403	0.1178	76	-0.20	001.07	0
001	Loss on Drying (%)	0226	9.380	0.1600	9.426	0.3403	0.1178	76	-0.14	001.07	0
001	Loss on Drying (%)	0682	9.390	0.0000	9.426	0.3403	0.1178	76	-0.11	001.07	0
001	Loss on Drying (%)	2009	9.393	0.0134	9.426	0.3403	0.1178	76	-0.10	001.07	0
001	Loss on Drying (%)	2109	9.410	0.0600	9.426	0.3403	0.1178	76	-0.05	001.07	0
001	Loss on Drying (%)	0868	9.420	0.0200	9.426	0.3403	0.1178	76	-0.02	001.03	0
001	Loss on Drying (%)	2062	9.431	0.0000	9.426	0.3403	0.1178	76	0.01	001.03	0
001	Loss on Drying (%)	0407	9.437	0.1044	9.426	0.3403	0.1178	76	0.03	001.07	0
001	Loss on Drying (%)	2259	9.457	0.0810	9.426	0.3403	0.1178	76	0.09	001.07	0
001	Loss on Drying (%)	0560	9.480	0.5200	9.426	0.3403	0.1178	76	0.16	001.99	0
001	Loss on Drying (%)	0089	9.510	0.0000	9.426	0.3403	0.1178	76	0.25	001.07	0
001	Loss on Drying (%)	0037	9.545	0.0100	9.426	0.3403	0.1178	76	0.35	001.99	0
001	Loss on Drying (%)	2059	9.550	0.0000	9.426	0.3403	0.1178	76	0.36	001.07	0
001	Loss on Drying (%)	0171	9.555	0.0100	9.426	0.3403	0.1178	76	0.38	001.07	0
001	Loss on Drying (%)	0619	9.560	0.0600	9.426	0.3403	0.1178	76	0.39	001.03	0
001	Loss on Drying (%)	0644	9.571	0.0290	9.426	0.3403	0.1178	76	0.42	001.05	0
001	Loss on Drying (%)	0003	9.575	0.0300	9.426	0.3403	0.1178	76	0.44	001.07	0
001	Loss on Drying (%)	0066	9.585	0.0700	9.426	0.3403	0.1178	76	0.47	001.07	0
001	Loss on Drying (%)	0297	9.585	0.1300	9.426	0.3403	0.1178	76	0.47	001.07	0
001	Loss on Drying (%)	0853	9.600	0.0000	9.426	0.3403	0.1178	76	0.51	001.07	0
001	Loss on Drying (%)	0035	9.600	0.0600	9.426	0.3403	0.1178	76	0.51	001.07	0
001	Loss on Drying (%)	2053	9.620	0.1800	9.426	0.3403	0.1178	76	0.57	001.07	0
001	Loss on Drying (%)	0186	9.625	0.2520	9.426	0.3403	0.1178	76	0.58	001.99	0
001	Loss on Drying (%)	0278	9.640	0.0200	9.426	0.3403	0.1178	76	0.63	001.07	0
001	Loss on Drying (%)	0910	9.650	0.1000	9.426	0.3403	0.1178	76	0.66	001.99	0
001	Loss on Drying (%)	0656	9.650	0.0000	9.426	0.3403	0.1178	76	0.66	001.99	0
001	Loss on Drying (%)	0581	9.660	0.0800	9.426	0.3403	0.1178	76	0.69	001.07	0
001	Loss on Drying (%)	0505	9.665	0.1300	9.426	0.3403	0.1178	76	0.70	001.07	0
001	Loss on Drying (%)	0309	9.682	0.0670	9.426	0.3403	0.1178	76	0.75	001.00	0
001	Loss on Drying (%)	0045	9.685	0.0100	9.426	0.3403	0.1178	76	0.76	001.07	0
001	Loss on Drying (%)	0028	9.685	0.0900	9.426	0.3403	0.1178	76	0.76	001.07	0
001	Loss on Drying (%)	0689	9.700	0.0000	9.426	0.3403	0.1178	76	0.80	001.07	0
001	Loss on Drying (%)	0893	9.710	0.0000	9.426	0.3403	0.1178	76	0.83	001.03	0
001	Loss on Drying (%)	0872	9.725	0.0700	9.426	0.3403	0.1178	76	0.88	001.07	0
001	Loss on Drying (%)	0074	9.750	0.3400	9.426	0.3403	0.1178	76	0.95	001.07	0
001	Loss on Drying (%)	2192	9.765	0.0300	9.426	0.3403	0.1178	76	1.00	001.07	0
001	Loss on Drying (%)	0693	9.906	0.4530	9.426	0.3403	0.1178	76	1.41	001.07	0
001	Loss on Drying (%)	0968	9.925	0.0100	9.426	0.3403	0.1178	76	1.47	001.05	0
001	Loss on Drying (%)	0098	9.945	0.0700	9.426	0.3403	0.1178	76	1.52	001.07	0
001	Loss on Drying (%)	0357	10.06	0.1222	9.426	0.3403	0.1178	76	1.87	001.99	0
001	Loss on Drying (%)	0843	10.09	0.4400	9.426	0.3403	0.1178	76	1.95	001.03	0
001	Loss on Drying (%)	0843	10.09	0.4400	9.426	0.3403	0.1178	76	1.95	001.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
001	Loss on Drying (%)	0019	10.18	0.0500	9.426	0.3403	0.1178	76	2.20	001.07	0
001	Loss on Drying (%)	2304	10.25	0.1200	9.426	0.3403	0.1178	76	2.42	001.99	0
002	Protein, Crude (%)	0011	13.19	0.1000	16.07	0.2604	0.1294	166	-11.06	002.06	0
002	Protein, Crude (%)	0889	14.84	0.1600	16.07	0.2604	0.1294	166	-4.74	002.11	0
002	Protein, Crude (%)	0852	15.24	0.1700	16.07	0.2604	0.1294	166	-3.21	002.05	0
002	Protein, Crude (%)	0893	15.30	0.0000	16.07	0.2604	0.1294	166	-2.96	002.05	0
002	Protein, Crude (%)	0644	15.31	0.0540	16.07	0.2604	0.1294	166	-2.91	002.06	0
002	Protein, Crude (%)	2291	15.44	0.1200	16.07	0.2604	0.1294	166	-2.42	002.05	0
002	Protein, Crude (%)	0853	15.45	0.5000	16.07	0.2604	0.1294	166	-2.38	002.06	0
002	Protein, Crude (%)	2246	15.46	0.0800	16.07	0.2604	0.1294	166	-2.34	002.06	0
002	Protein, Crude (%)	0563	15.48	0.3205	16.07	0.2604	0.1294	166	-2.26	002.08	0
002	Protein, Crude (%)	0529	15.50	0.0100	16.07	0.2604	0.1294	166	-2.21	002.06	0
002	Protein, Crude (%)	0144	15.55	0.2800	16.07	0.2604	0.1294	166	-2.00	002.06	0
002	Protein, Crude (%)	0910	15.65	0.5000	16.07	0.2604	0.1294	166	-1.61	002.01	0
002	Protein, Crude (%)	2302	15.66	0.2500	16.07	0.2604	0.1294	166	-1.59	002.06	0
002	Protein, Crude (%)	2295	15.68	0.0100	16.07	0.2604	0.1294	166	-1.52	002.05	0
002	Protein, Crude (%)	0589	15.70	0.2300	16.07	0.2604	0.1294	166	-1.44	002.06	0
002	Protein, Crude (%)	0598	15.70	0.4100	16.07	0.2604	0.1294	166	-1.44	002.06	0
002	Protein, Crude (%)	0345	15.70	0.2000	16.07	0.2604	0.1294	166	-1.42	002.06	0
002	Protein, Crude (%)	0539	15.71	0.2900	16.07	0.2604	0.1294	166	-1.40	002.06	0
002	Protein, Crude (%)	0357	15.72	0.0625	16.07	0.2604	0.1294	166	-1.35	002.06	0
002	Protein, Crude (%)	0202	15.75	0.2400	16.07	0.2604	0.1294	166	-1.23	002.06	0
002	Protein, Crude (%)	0876	15.75	0.1000	16.07	0.2604	0.1294	166	-1.23	002.06	0
002	Protein, Crude (%)	2303	15.75	0.6200	16.07	0.2604	0.1294	166	-1.23	002.06	0
002	Protein, Crude (%)	0226	15.76	0.1100	16.07	0.2604	0.1294	166	-1.21	002.06	0
002	Protein, Crude (%)	2144	15.78	0.0400	16.07	0.2604	0.1294	166	-1.11	002.05	0
002	Protein, Crude (%)	0939	15.79	0.1100	16.07	0.2604	0.1294	166	-1.09	002.05	0
002	Protein, Crude (%)	0939	15.79	0.0800	16.07	0.2604	0.1294	166	-1.07	002.06	0
002	Protein, Crude (%)	0148	15.80	0.0920	16.07	0.2604	0.1294	166	-1.05	002.06	0
002	Protein, Crude (%)	0619	15.80	0.2000	16.07	0.2604	0.1294	166	-1.04	002.05	0
002	Protein, Crude (%)	0510	15.80	0.0000	16.07	0.2604	0.1294	166	-1.04	002.06	0
002	Protein, Crude (%)	0885	15.83	0.0354	16.07	0.2604	0.1294	166	-0.91	002.05	0
002	Protein, Crude (%)	0036	15.85	0.1921	16.07	0.2604	0.1294	166	-0.85	002.02	0
002	Protein, Crude (%)	0674	15.85	0.0000	16.07	0.2604	0.1294	166	-0.84	002.05	0
002	Protein, Crude (%)	2109	15.86	0.0800	16.07	0.2604	0.1294	166	-0.81	002.05	0
002	Protein, Crude (%)	0004	15.87	0.3800	16.07	0.2604	0.1294	166	-0.77	002.06	0
002	Protein, Crude (%)	0868	15.88	0.0800	16.07	0.2604	0.1294	166	-0.73	002.05	0
002	Protein, Crude (%)	0689	15.90	0.2000	16.07	0.2604	0.1294	166	-0.65	002.05	0
002	Protein, Crude (%)	0032	15.90	0.0400	16.07	0.2604	0.1294	166	-0.64	002.06	0
002	Protein, Crude (%)	0066	15.92	0.0900	16.07	0.2604	0.1294	166	-0.59	002.02	0
002	Protein, Crude (%)	2022	15.92	0.0100	16.07	0.2604	0.1294	166	-0.59	002.05	0
002	Protein, Crude (%)	0546	15.92	0.0660	16.07	0.2604	0.1294	166	-0.56	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0964	15.93	0.0010	16.07	0.2604	0.1294	166	-0.54	002.06	0
002	Protein, Crude (%)	2150	15.94	0.0100	16.07	0.2604	0.1294	166	-0.52	002.06	0
002	Protein, Crude (%)	0675	15.94	0.0500	16.07	0.2604	0.1294	166	-0.52	002.05	0
002	Protein, Crude (%)	0968	15.94	0.0000	16.07	0.2604	0.1294	166	-0.50	002.01	0
002	Protein, Crude (%)	0006	15.95	0.1500	16.07	0.2604	0.1294	166	-0.48	002.06	0
002	Protein, Crude (%)	0652	15.95	0.1000	16.07	0.2604	0.1294	166	-0.46	002.01	0
002	Protein, Crude (%)	0171	15.95	0.0400	16.07	0.2604	0.1294	166	-0.46	002.06	0
002	Protein, Crude (%)	0619	15.95	0.1000	16.07	0.2604	0.1294	166	-0.46	002.06	0
002	Protein, Crude (%)	0263	15.96	0.0620	16.07	0.2604	0.1294	166	-0.42	002.06	0
002	Protein, Crude (%)	0881	15.96	0.0750	16.07	0.2604	0.1294	166	-0.41	002.05	0
002	Protein, Crude (%)	0693	15.97	0.1980	16.07	0.2604	0.1294	166	-0.37	002.06	0
002	Protein, Crude (%)	0581	15.98	0.2100	16.07	0.2604	0.1294	166	-0.36	002.06	0
002	Protein, Crude (%)	0953	15.98	0.5700	16.07	0.2604	0.1294	166	-0.36	002.06	0
002	Protein, Crude (%)	2129	15.98	0.0760	16.07	0.2604	0.1294	166	-0.36	002.06	0
002	Protein, Crude (%)	0870	15.98	0.0160	16.07	0.2604	0.1294	166	-0.36	002.01	0
002	Protein, Crude (%)	0028	15.98	0.0200	16.07	0.2604	0.1294	166	-0.34	002.06	0
002	Protein, Crude (%)	0003	15.99	0.0800	16.07	0.2604	0.1294	166	-0.31	002.06	0
002	Protein, Crude (%)	0662	16.00	0.1250	16.07	0.2604	0.1294	166	-0.27	002.01	0
002	Protein, Crude (%)	0014	16.00	0.0000	16.07	0.2604	0.1294	166	-0.27	002.06	0
002	Protein, Crude (%)	2109	16.00	0.0600	16.07	0.2604	0.1294	166	-0.27	002.06	0
002	Protein, Crude (%)	0728	16.01	0.1500	16.07	0.2604	0.1294	166	-0.25	002.04	0
002	Protein, Crude (%)	2184	16.01	0.0700	16.07	0.2604	0.1294	166	-0.25	002.06	0
002	Protein, Crude (%)	0857	16.01	0.1000	16.07	0.2604	0.1294	166	-0.23	002.06	0
002	Protein, Crude (%)	2196	16.01	0.0000	16.07	0.2604	0.1294	166	-0.23	002.01	0
002	Protein, Crude (%)	0878	16.01	0.0000	16.07	0.2604	0.1294	166	-0.23	002.05	0
002	Protein, Crude (%)	0354	16.02	0.0100	16.07	0.2604	0.1294	166	-0.21	002.05	0
002	Protein, Crude (%)	0968	16.02	0.0000	16.07	0.2604	0.1294	166	-0.19	002.06	0
002	Protein, Crude (%)	2168	16.02	0.0200	16.07	0.2604	0.1294	166	-0.19	002.06	0
002	Protein, Crude (%)	0297	16.03	0.0500	16.07	0.2604	0.1294	166	-0.17	002.06	0
002	Protein, Crude (%)	0353	16.03	0.0700	16.07	0.2604	0.1294	166	-0.17	002.06	0
002	Protein, Crude (%)	0164	16.03	0.0600	16.07	0.2604	0.1294	166	-0.15	002.01	0
002	Protein, Crude (%)	0164	16.03	0.0600	16.07	0.2604	0.1294	166	-0.15	002.06	0
002	Protein, Crude (%)	2009	16.03	0.0130	16.07	0.2604	0.1294	166	-0.14	002.05	0
002	Protein, Crude (%)	0676	16.04	0.1900	16.07	0.2604	0.1294	166	-0.13	002.06	0
002	Protein, Crude (%)	0505	16.04	0.0400	16.07	0.2604	0.1294	166	-0.11	002.06	0
002	Protein, Crude (%)	0948	16.05	0.0900	16.07	0.2604	0.1294	166	-0.09	002.05	0
002	Protein, Crude (%)	0035	16.05	0.1100	16.07	0.2604	0.1294	166	-0.09	002.06	0
002	Protein, Crude (%)	0870	16.05	0.0768	16.07	0.2604	0.1294	166	-0.08	002.06	0
002	Protein, Crude (%)	2181	16.05	0.0600	16.07	0.2604	0.1294	166	-0.08	002.05	0
002	Protein, Crude (%)	0098	16.05	0.3000	16.07	0.2604	0.1294	166	-0.08	002.06	0
002	Protein, Crude (%)	2192	16.06	0.0100	16.07	0.2604	0.1294	166	-0.06	002.05	0
002	Protein, Crude (%)	0108	16.07	0.0100	16.07	0.2604	0.1294	166	-0.02	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0309	16.07	0.0500	16.07	0.2604	0.1294	166	-0.02	002.06	0
002	Protein, Crude (%)	0018	16.08	0.1500	16.07	0.2604	0.1294	166	0.02	002.06	0
002	Protein, Crude (%)	0358	16.08	0.0500	16.07	0.2604	0.1294	166	0.02	002.06	0
002	Protein, Crude (%)	2062	16.08	0.0257	16.07	0.2604	0.1294	166	0.04	002.05	0
002	Protein, Crude (%)	0859	16.08	0.2570	16.07	0.2604	0.1294	166	0.05	002.06	0
002	Protein, Crude (%)	2146	16.09	0.0500	16.07	0.2604	0.1294	166	0.06	002.05	0
002	Protein, Crude (%)	0626	16.09	0.0500	16.07	0.2604	0.1294	166	0.06	002.06	0
002	Protein, Crude (%)	0098	16.09	0.1400	16.07	0.2604	0.1294	166	0.08	002.08	0
002	Protein, Crude (%)	0123	16.10	0.0500	16.07	0.2604	0.1294	166	0.10	002.06	0
002	Protein, Crude (%)	0354	16.10	0.0300	16.07	0.2604	0.1294	166	0.10	002.06	0
002	Protein, Crude (%)	0278	16.10	0.0000	16.07	0.2604	0.1294	166	0.12	002.06	0
002	Protein, Crude (%)	0687	16.10	0.2000	16.07	0.2604	0.1294	166	0.12	002.06	0
002	Protein, Crude (%)	2053	16.10	0.2000	16.07	0.2604	0.1294	166	0.12	002.06	0
002	Protein, Crude (%)	0169	16.11	0.0200	16.07	0.2604	0.1294	166	0.15	002.00	0
002	Protein, Crude (%)	0536	16.12	0.2300	16.07	0.2604	0.1294	166	0.17	002.05	0
002	Protein, Crude (%)	2012	16.12	0.0900	16.07	0.2604	0.1294	166	0.17	002.05	0
002	Protein, Crude (%)	0142	16.12	0.1000	16.07	0.2604	0.1294	166	0.19	002.06	0
002	Protein, Crude (%)	0590	16.12	0.1000	16.07	0.2604	0.1294	166	0.19	002.06	0
002	Protein, Crude (%)	2081	16.13	0.1400	16.07	0.2604	0.1294	166	0.23	002.06	0
002	Protein, Crude (%)	0089	16.14	0.0000	16.07	0.2604	0.1294	166	0.27	002.06	0
002	Protein, Crude (%)	0242	16.14	0.0200	16.07	0.2604	0.1294	166	0.27	002.06	0
002	Protein, Crude (%)	2073	16.15	0.0100	16.07	0.2604	0.1294	166	0.29	002.05	0
002	Protein, Crude (%)	0407	16.15	0.0300	16.07	0.2604	0.1294	166	0.30	002.06	0
002	Protein, Crude (%)	0683	16.15	0.0400	16.07	0.2604	0.1294	166	0.31	002.05	0
002	Protein, Crude (%)	0045	16.15	0.1000	16.07	0.2604	0.1294	166	0.31	002.06	0
002	Protein, Crude (%)	0175	16.15	0.1000	16.07	0.2604	0.1294	166	0.31	002.06	0
002	Protein, Crude (%)	0083	16.16	0.0000	16.07	0.2604	0.1294	166	0.35	002.06	0
002	Protein, Crude (%)	2213	16.16	0.1800	16.07	0.2604	0.1294	166	0.35	002.06	0
002	Protein, Crude (%)	0298	16.17	0.0200	16.07	0.2604	0.1294	166	0.39	002.06	0
002	Protein, Crude (%)	0726	16.17	0.1200	16.07	0.2604	0.1294	166	0.39	002.06	0
002	Protein, Crude (%)	2216	16.17	0.1400	16.07	0.2604	0.1294	166	0.39	002.06	0
002	Protein, Crude (%)	0574	16.18	0.1500	16.07	0.2604	0.1294	166	0.40	002.06	0
002	Protein, Crude (%)	0904	16.18	0.0400	16.07	0.2604	0.1294	166	0.42	002.06	0
002	Protein, Crude (%)	0897	16.19	0.0200	16.07	0.2604	0.1294	166	0.46	002.05	0
002	Protein, Crude (%)	0100	16.19	0.2000	16.07	0.2604	0.1294	166	0.46	002.06	0
002	Protein, Crude (%)	0848	16.19	0.0400	16.07	0.2604	0.1294	166	0.46	002.06	0
002	Protein, Crude (%)	2188	16.20	0.0000	16.07	0.2604	0.1294	166	0.50	002.01	0
002	Protein, Crude (%)	0294	16.20	0.0000	16.07	0.2604	0.1294	166	0.50	002.06	0
002	Protein, Crude (%)	0682	16.20	0.0000	16.07	0.2604	0.1294	166	0.50	002.06	0
002	Protein, Crude (%)	0956	16.20	0.2000	16.07	0.2604	0.1294	166	0.50	002.06	0
002	Protein, Crude (%)	0001	16.21	0.3600	16.07	0.2604	0.1294	166	0.54	002.06	0
002	Protein, Crude (%)	0019	16.21	0.0200	16.07	0.2604	0.1294	166	0.54	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0009	16.22	0.1500	16.07	0.2604	0.1294	166	0.56	002.06	0
002	Protein, Crude (%)	0037	16.22	0.4300	16.07	0.2604	0.1294	166	0.56	002.06	0
002	Protein, Crude (%)	2006	16.22	0.2200	16.07	0.2604	0.1294	166	0.58	002.05	0
002	Protein, Crude (%)	0190	16.23	0.0500	16.07	0.2604	0.1294	166	0.60	002.06	0
002	Protein, Crude (%)	0868	16.23	0.0100	16.07	0.2604	0.1294	166	0.60	002.06	0
002	Protein, Crude (%)	0541	16.23	0.0400	16.07	0.2604	0.1294	166	0.62	002.06	0
002	Protein, Crude (%)	0675	16.24	0.0300	16.07	0.2604	0.1294	166	0.64	002.06	0
002	Protein, Crude (%)	0872	16.24	0.3000	16.07	0.2604	0.1294	166	0.65	002.06	0
002	Protein, Crude (%)	0010	16.25	0.1000	16.07	0.2604	0.1294	166	0.69	002.06	0
002	Protein, Crude (%)	0265	16.25	0.3000	16.07	0.2604	0.1294	166	0.69	002.06	0
002	Protein, Crude (%)	0910	16.25	0.1000	16.07	0.2604	0.1294	166	0.69	002.06	0
002	Protein, Crude (%)	2174	16.26	0.0700	16.07	0.2604	0.1294	166	0.71	002.06	0
002	Protein, Crude (%)	2302	16.26	0.9900	16.07	0.2604	0.1294	166	0.71	002.11	0
002	Protein, Crude (%)	0036	16.28	0.2400	16.07	0.2604	0.1294	166	0.81	002.06	0
002	Protein, Crude (%)	0553	16.28	0.0990	16.07	0.2604	0.1294	166	0.81	002.06	0
002	Protein, Crude (%)	0914	16.29	0.0900	16.07	0.2604	0.1294	166	0.83	002.06	0
002	Protein, Crude (%)	0034	16.29	0.0610	16.07	0.2604	0.1294	166	0.84	002.06	0
002	Protein, Crude (%)	0425	16.31	0.0000	16.07	0.2604	0.1294	166	0.92	002.06	0
002	Protein, Crude (%)	0889	16.33	0.0600	16.07	0.2604	0.1294	166	1.00	002.01	0
002	Protein, Crude (%)	0571	16.39	0.0500	16.07	0.2604	0.1294	166	1.21	002.06	0
002	Protein, Crude (%)	0918	16.39	0.0900	16.07	0.2604	0.1294	166	1.21	002.06	0
002	Protein, Crude (%)	0643	16.47	0.5500	16.07	0.2604	0.1294	166	1.52	002.99	0
002	Protein, Crude (%)	2246	16.49	0.1500	16.07	0.2604	0.1294	166	1.60	002.05	0
002	Protein, Crude (%)	0852	16.54	0.0300	16.07	0.2604	0.1294	166	1.79	002.11	0
002	Protein, Crude (%)	0229	16.55	0.3300	16.07	0.2604	0.1294	166	1.83	002.06	0
002	Protein, Crude (%)	0618	16.55	0.0370	16.07	0.2604	0.1294	166	1.84	002.06	0
002	Protein, Crude (%)	0042	16.55	0.3000	16.07	0.2604	0.1294	166	1.84	002.06	0
002	Protein, Crude (%)	0520	16.55	0.5000	16.07	0.2604	0.1294	166	1.84	002.06	0
002	Protein, Crude (%)	2059	16.56	0.0844	16.07	0.2604	0.1294	166	1.87	002.06	0
002	Protein, Crude (%)	0843	16.57	0.0100	16.07	0.2604	0.1294	166	1.90	002.06	0
002	Protein, Crude (%)	0692	16.65	0.7000	16.07	0.2604	0.1294	166	2.23	002.06	0
002	Protein, Crude (%)	0590	16.65	0.1000	16.07	0.2604	0.1294	166	2.23	002.11	0
002	Protein, Crude (%)	0186	16.68	0.1900	16.07	0.2604	0.1294	166	2.32	002.06	0
002	Protein, Crude (%)	2012	16.69	0.0000	16.07	0.2604	0.1294	166	2.38	002.06	0
002	Protein, Crude (%)	0553	16.74	0.0000	16.07	0.2604	0.1294	166	2.57	002.11	0
002	Protein, Crude (%)	0656	17.01	0.0200	16.07	0.2604	0.1294	166	3.61	002.06	0
002	Protein, Crude (%)	0297	17.16	0.0000	16.07	0.2604	0.1294	166	4.19	002.11	0
002	Protein, Crude (%)	2155	17.48	0.7890	16.07	0.2604	0.1294	166	5.43	002.06	0
002	Protein, Crude (%)	0148	17.50	0.0000	16.07	0.2604	0.1294	166	5.49	002.11	0
002	Protein, Crude (%)	2259	18.14	0.0370	16.07	0.2604	0.1294	166	7.93	002.01	0
002	Protein, Crude (%)	2178	18.44	0.5100	16.07	0.2604	0.1294	166	9.08	002.06	0
002	Protein, Crude (%)	0011	18.64	0.0000	16.07	0.2604	0.1294	166	9.87	002.11	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	2076	11.32	9.926	16.07	0.2604	0.1294	166	-18.25	002.06	2
002	Protein, Crude (%)	2268	20.21	0.0900	16.07	0.2604	0.1294	166	15.88	002.06	2
003	Fat, Crude (%)	0357	0.9500	0.1000	2.502	0.4166	0.0955	137	-3.73	003.14	0
003	Fat, Crude (%)	0175	1.435	0.0300	2.502	0.4166	0.0955	137	-2.56	003.14	0
003	Fat, Crude (%)	2129	1.447	0.1109	2.502	0.4166	0.0955	137	-2.53	003.14	0
003	Fat, Crude (%)	0520	1.635	0.0500	2.502	0.4166	0.0955	137	-2.08	003.14	0
003	Fat, Crude (%)	2302	1.665	0.1300	2.502	0.4166	0.0955	137	-2.01	003.14	0
003	Fat, Crude (%)	0893	1.730	0.0000	2.502	0.4166	0.0955	137	-1.85	003.12	0
003	Fat, Crude (%)	0186	1.755	0.0300	2.502	0.4166	0.0955	137	-1.79	003.14	0
003	Fat, Crude (%)	0265	1.795	0.0100	2.502	0.4166	0.0955	137	-1.70	003.14	0
003	Fat, Crude (%)	0019	1.825	0.0300	2.502	0.4166	0.0955	137	-1.63	003.14	0
003	Fat, Crude (%)	0536	1.845	0.2900	2.502	0.4166	0.0955	137	-1.58	003.99	0
003	Fat, Crude (%)	0089	1.900	0.0000	2.502	0.4166	0.0955	137	-1.45	003.14	0
003	Fat, Crude (%)	2076	1.904	0.1823	2.502	0.4166	0.0955	137	-1.44	003.14	0
003	Fat, Crude (%)	0939	1.940	0.4800	2.502	0.4166	0.0955	137	-1.35	003.14	0
003	Fat, Crude (%)	2073	1.945	0.0300	2.502	0.4166	0.0955	137	-1.34	003.10	0
003	Fat, Crude (%)	0358	1.945	0.0500	2.502	0.4166	0.0955	137	-1.34	003.14	0
003	Fat, Crude (%)	0013	1.965	0.2900	2.502	0.4166	0.0955	137	-1.29	003.14	0
003	Fat, Crude (%)	2062	1.973	0.0060	2.502	0.4166	0.0955	137	-1.27	003.12	0
003	Fat, Crude (%)	0505	1.975	0.0300	2.502	0.4166	0.0955	137	-1.27	003.14	0
003	Fat, Crude (%)	0083	1.995	0.0700	2.502	0.4166	0.0955	137	-1.22	003.14	0
003	Fat, Crude (%)	0956	2.000	0.2000	2.502	0.4166	0.0955	137	-1.21	003.14	0
003	Fat, Crude (%)	0968	2.010	0.1600	2.502	0.4166	0.0955	137	-1.18	003.14	0
003	Fat, Crude (%)	0034	2.050	0.2200	2.502	0.4166	0.0955	137	-1.09	003.14	0
003	Fat, Crude (%)	0144	2.060	0.2400	2.502	0.4166	0.0955	137	-1.06	003.14	0
003	Fat, Crude (%)	0108	2.090	0.1000	2.502	0.4166	0.0955	137	-0.99	003.14	0
003	Fat, Crude (%)	0529	2.095	0.0700	2.502	0.4166	0.0955	137	-0.98	003.14	0
003	Fat, Crude (%)	0581	2.095	0.0300	2.502	0.4166	0.0955	137	-0.98	003.14	0
003	Fat, Crude (%)	0035	2.135	0.0300	2.502	0.4166	0.0955	137	-0.88	003.14	0
003	Fat, Crude (%)	0229	2.150	0.1000	2.502	0.4166	0.0955	137	-0.85	003.14	0
003	Fat, Crude (%)	2022	2.160	0.0200	2.502	0.4166	0.0955	137	-0.82	003.10	0
003	Fat, Crude (%)	0004	2.165	0.0300	2.502	0.4166	0.0955	137	-0.81	003.14	0
003	Fat, Crude (%)	0263	2.182	0.0200	2.502	0.4166	0.0955	137	-0.77	003.10	0
003	Fat, Crude (%)	0148	2.200	0.0200	2.502	0.4166	0.0955	137	-0.73	003.14	0
003	Fat, Crude (%)	0682	2.200	0.0000	2.502	0.4166	0.0955	137	-0.73	003.14	0
003	Fat, Crude (%)	2246	2.210	0.0600	2.502	0.4166	0.0955	137	-0.70	003.13	0
003	Fat, Crude (%)	0298	2.220	0.0200	2.502	0.4166	0.0955	137	-0.68	003.14	0
003	Fat, Crude (%)	0037	2.230	0.0200	2.502	0.4166	0.0955	137	-0.65	003.14	0
003	Fat, Crude (%)	0202	2.235	0.0100	2.502	0.4166	0.0955	137	-0.64	003.14	0
003	Fat, Crude (%)	0619	2.280	0.0400	2.502	0.4166	0.0955	137	-0.53	003.10	0
003	Fat, Crude (%)	0546	2.290	0.0200	2.502	0.4166	0.0955	137	-0.51	003.99	0
003	Fat, Crude (%)	0878	2.295	0.1100	2.502	0.4166	0.0955	137	-0.50	003.14	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat, Crude (%)	0563	2.322	0.1052	2.502	0.4166	0.0955	137	-0.43	003.01	0
003	Fat, Crude (%)	0563	2.322	0.1052	2.502	0.4166	0.0955	137	-0.43	003.14	0
003	Fat, Crude (%)	0897	2.325	0.0100	2.502	0.4166	0.0955	137	-0.43	003.10	0
003	Fat, Crude (%)	0345	2.340	0.0200	2.502	0.4166	0.0955	137	-0.39	003.00	0
003	Fat, Crude (%)	0853	2.350	0.1000	2.502	0.4166	0.0955	137	-0.37	003.14	0
003	Fat, Crude (%)	0407	2.359	0.2595	2.502	0.4166	0.0955	137	-0.35	003.14	0
003	Fat, Crude (%)	0868	2.360	0.0200	2.502	0.4166	0.0955	137	-0.34	003.14	0
003	Fat, Crude (%)	2295	2.365	0.0300	2.502	0.4166	0.0955	137	-0.33	003.06	0
003	Fat, Crude (%)	2181	2.370	0.1000	2.502	0.4166	0.0955	137	-0.32	003.06	0
003	Fat, Crude (%)	0009	2.385	0.0700	2.502	0.4166	0.0955	137	-0.28	003.14	0
003	Fat, Crude (%)	0918	2.390	0.0600	2.502	0.4166	0.0955	137	-0.27	003.06	0
003	Fat, Crude (%)	0027	2.402	0.0960	2.502	0.4166	0.0955	137	-0.24	003.09	0
003	Fat, Crude (%)	2268	2.405	0.0500	2.502	0.4166	0.0955	137	-0.23	003.06	0
003	Fat, Crude (%)	0229	2.410	0.5600	2.502	0.4166	0.0955	137	-0.22	003.06	0
003	Fat, Crude (%)	0728	2.415	0.0300	2.502	0.4166	0.0955	137	-0.21	003.10	0
003	Fat, Crude (%)	2155	2.422	0.5300	2.502	0.4166	0.0955	137	-0.19	003.10	0
003	Fat, Crude (%)	0036	2.422	0.0420	2.502	0.4166	0.0955	137	-0.19	003.10	0
003	Fat, Crude (%)	0001	2.431	0.0660	2.502	0.4166	0.0955	137	-0.17	003.14	0
003	Fat, Crude (%)	0598	2.440	0.0200	2.502	0.4166	0.0955	137	-0.15	003.14	0
003	Fat, Crude (%)	0013	2.445	0.1900	2.502	0.4166	0.0955	137	-0.14	003.09	0
003	Fat, Crude (%)	0171	2.465	0.0700	2.502	0.4166	0.0955	137	-0.09	003.12	0
003	Fat, Crude (%)	0035	2.470	0.0600	2.502	0.4166	0.0955	137	-0.08	003.00	0
003	Fat, Crude (%)	2302	2.470	0.4800	2.502	0.4166	0.0955	137	-0.08	003.11	0
003	Fat, Crude (%)	2196	2.470	0.0000	2.502	0.4166	0.0955	137	-0.08	003.10	0
003	Fat, Crude (%)	0889	2.473	0.0050	2.502	0.4166	0.0955	137	-0.07	003.11	0
003	Fat, Crude (%)	0683	2.475	0.1500	2.502	0.4166	0.0955	137	-0.07	003.14	0
003	Fat, Crude (%)	0885	2.476	0.0762	2.502	0.4166	0.0955	137	-0.06	003.00	0
003	Fat, Crude (%)	2192	2.485	0.0100	2.502	0.4166	0.0955	137	-0.04	003.00	0
003	Fat, Crude (%)	0242	2.495	0.0500	2.502	0.4166	0.0955	137	-0.02	003.10	0
003	Fat, Crude (%)	2188	2.500	0.0400	2.502	0.4166	0.0955	137	-0.01	003.10	0
003	Fat, Crude (%)	0309	2.502	0.0411	2.502	0.4166	0.0955	137	0.00	003.00	0
003	Fat, Crude (%)	0226	2.535	0.3300	2.502	0.4166	0.0955	137	0.08	003.09	0
003	Fat, Crude (%)	0098	2.560	0.0600	2.502	0.4166	0.0955	137	0.14	003.10	0
003	Fat, Crude (%)	0881	2.564	0.2370	2.502	0.4166	0.0955	137	0.15	003.10	0
003	Fat, Crude (%)	2009	2.579	0.0182	2.502	0.4166	0.0955	137	0.18	003.10	0
003	Fat, Crude (%)	0683	2.590	0.0600	2.502	0.4166	0.0955	137	0.21	003.00	0
003	Fat, Crude (%)	2291	2.595	0.2700	2.502	0.4166	0.0955	137	0.22	003.13	0
003	Fat, Crude (%)	0689	2.600	0.0000	2.502	0.4166	0.0955	137	0.23	003.10	0
003	Fat, Crude (%)	0003	2.615	0.5100	2.502	0.4166	0.0955	137	0.27	003.14	0
003	Fat, Crude (%)	0123	2.620	0.0000	2.502	0.4166	0.0955	137	0.28	003.13	0
003	Fat, Crude (%)	0066	2.620	0.2000	2.502	0.4166	0.0955	137	0.28	003.14	0
003	Fat, Crude (%)	0169	2.630	0.2600	2.502	0.4166	0.0955	137	0.31	003.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat, Crude (%)	0045	2.650	0.1000	2.502	0.4166	0.0955	137	0.35	003.10	0
003	Fat, Crude (%)	0062	2.655	0.0040	2.502	0.4166	0.0955	137	0.37	003.10	0
003	Fat, Crude (%)	0693	2.657	0.0840	2.502	0.4166	0.0955	137	0.37	003.10	0
003	Fat, Crude (%)	0042	2.660	0.0800	2.502	0.4166	0.0955	137	0.38	003.10	0
003	Fat, Crude (%)	2184	2.660	0.0800	2.502	0.4166	0.0955	137	0.38	003.10	0
003	Fat, Crude (%)	0028	2.660	0.1000	2.502	0.4166	0.0955	137	0.38	003.13	0
003	Fat, Crude (%)	0541	2.665	0.1700	2.502	0.4166	0.0955	137	0.39	003.10	0
003	Fat, Crude (%)	0164	2.675	0.1700	2.502	0.4166	0.0955	137	0.41	003.01	0
003	Fat, Crude (%)	0687	2.700	0.2000	2.502	0.4166	0.0955	137	0.47	003.06	0
003	Fat, Crude (%)	0689	2.700	0.0000	2.502	0.4166	0.0955	137	0.47	003.06	0
003	Fat, Crude (%)	2006	2.730	0.1000	2.502	0.4166	0.0955	137	0.55	003.13	0
003	Fat, Crude (%)	0098	2.740	0.0000	2.502	0.4166	0.0955	137	0.57	003.13	0
003	Fat, Crude (%)	0619	2.740	0.0400	2.502	0.4166	0.0955	137	0.57	003.14	0
003	Fat, Crude (%)	0889	2.750	0.0800	2.502	0.4166	0.0955	137	0.59	003.10	0
003	Fat, Crude (%)	0278	2.750	0.1000	2.502	0.4166	0.0955	137	0.59	003.14	0
003	Fat, Crude (%)	0297	2.770	0.0000	2.502	0.4166	0.0955	137	0.64	003.06	0
003	Fat, Crude (%)	2012	2.770	0.0400	2.502	0.4166	0.0955	137	0.64	003.09	0
003	Fat, Crude (%)	0190	2.770	0.1200	2.502	0.4166	0.0955	137	0.64	003.14	0
003	Fat, Crude (%)	0682	2.790	0.0000	2.502	0.4166	0.0955	137	0.69	003.06	0
003	Fat, Crude (%)	0004	2.795	0.0500	2.502	0.4166	0.0955	137	0.70	003.09	0
003	Fat, Crude (%)	0098	2.795	0.1500	2.502	0.4166	0.0955	137	0.70	003.09	0
003	Fat, Crude (%)	0148	2.800	0.0000	2.502	0.4166	0.0955	137	0.71	003.11	0
003	Fat, Crude (%)	0644	2.802	0.0900	2.502	0.4166	0.0955	137	0.72	003.06	0
003	Fat, Crude (%)	0870	2.804	0.1590	2.502	0.4166	0.0955	137	0.72	003.09	0
003	Fat, Crude (%)	0870	2.804	0.1590	2.502	0.4166	0.0955	137	0.72	003.10	0
003	Fat, Crude (%)	0870	2.804	0.1590	2.502	0.4166	0.0955	137	0.72	003.13	0
003	Fat, Crude (%)	0294	2.805	0.0100	2.502	0.4166	0.0955	137	0.73	003.06	0
003	Fat, Crude (%)	0100	2.805	0.1100	2.502	0.4166	0.0955	137	0.73	003.10	0
003	Fat, Crude (%)	0171	2.825	0.3100	2.502	0.4166	0.0955	137	0.77	003.14	0
003	Fat, Crude (%)	2053	2.830	0.0800	2.502	0.4166	0.0955	137	0.79	003.00	0
003	Fat, Crude (%)	0148	2.840	0.2400	2.502	0.4166	0.0955	137	0.81	003.10	0
003	Fat, Crude (%)	0941	2.850	0.2000	2.502	0.4166	0.0955	137	0.83	003.99	0
003	Fat, Crude (%)	0726	2.855	0.0500	2.502	0.4166	0.0955	137	0.85	003.14	0
003	Fat, Crude (%)	0590	2.865	0.0700	2.502	0.4166	0.0955	137	0.87	003.11	0
003	Fat, Crude (%)	0964	2.895	0.0340	2.502	0.4166	0.0955	137	0.94	003.09	0
003	Fat, Crude (%)	0354	2.940	0.0200	2.502	0.4166	0.0955	137	1.05	003.09	0
003	Fat, Crude (%)	0948	2.945	0.0100	2.502	0.4166	0.0955	137	1.06	003.14	0
003	Fat, Crude (%)	0425	2.960	0.0000	2.502	0.4166	0.0955	137	1.10	003.06	0
003	Fat, Crude (%)	0674	2.960	0.0000	2.502	0.4166	0.0955	137	1.10	003.09	0
003	Fat, Crude (%)	0910	2.980	0.0600	2.502	0.4166	0.0955	137	1.15	003.06	0
003	Fat, Crude (%)	0948	3.005	0.0300	2.502	0.4166	0.0955	137	1.21	003.09	0
003	Fat, Crude (%)	0353	3.010	0.1800	2.502	0.4166	0.0955	137	1.22	003.09	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
003	Fat, Crude (%)	0590	3.010	0.1200	2.502	0.4166	0.0955	137	1.22	003.09	0
003	Fat, Crude (%)	0032	3.030	0.0400	2.502	0.4166	0.0955	137	1.27	003.00	0
003	Fat, Crude (%)	0553	3.040	0.0200	2.502	0.4166	0.0955	137	1.29	003.11	0
003	Fat, Crude (%)	0510	3.050	0.1000	2.502	0.4166	0.0955	137	1.31	003.09	0
003	Fat, Crude (%)	0175	3.080	0.0400	2.502	0.4166	0.0955	137	1.39	003.00	0
003	Fat, Crude (%)	2109	3.135	0.0300	2.502	0.4166	0.0955	137	1.52	003.00	0
003	Fat, Crude (%)	0675	3.195	0.0100	2.502	0.4166	0.0955	137	1.66	003.09	0
003	Fat, Crude (%)	0297	3.345	0.0700	2.502	0.4166	0.0955	137	2.02	003.11	0
003	Fat, Crude (%)	0656	3.390	0.0000	2.502	0.4166	0.0955	137	2.13	003.99	0
003	Fat, Crude (%)	0852	3.440	0.0200	2.502	0.4166	0.0955	137	2.25	003.11	0
003	Fat, Crude (%)	0574	3.450	0.0000	2.502	0.4166	0.0955	137	2.27	003.14	0
003	Fat, Crude (%)	2144	3.960	0.0600	2.502	0.4166	0.0955	137	3.50	003.14	0
003	Fat, Crude (%)	2144	4.080	0.1000	2.502	0.4166	0.0955	137	3.79	003.00	0
003	Fat, Crude (%)	0692	1.700	0.8000	2.502	0.4166	0.0955	137	-1.93	003.14	1
003	Fat, Crude (%)	0876	2.350	0.9000	2.502	0.4166	0.0955	137	-0.37	003.14	1
003	Fat, Crude (%)	0011	6.250	0.0000	2.502	0.4166	0.0955	137	9.00	003.11	2
004	Fiber, Crude (%)	0682	2.200	0.0000	3.725	0.3718	0.1250	109	-4.10	004.07	0
004	Fiber, Crude (%)	0510	2.450	0.1000	3.725	0.3718	0.1250	109	-3.43	004.07	0
004	Fiber, Crude (%)	0265	2.595	0.0700	3.725	0.3718	0.1250	109	-3.04	004.07	0
004	Fiber, Crude (%)	2146	2.900	0.0200	3.725	0.3718	0.1250	109	-2.22	004.00	0
004	Fiber, Crude (%)	2012	2.985	0.0700	3.725	0.3718	0.1250	109	-1.99	004.07	0
004	Fiber, Crude (%)	0656	3.030	0.0000	3.725	0.3718	0.1250	109	-1.87	004.07	0
004	Fiber, Crude (%)	0354	3.075	0.0300	3.725	0.3718	0.1250	109	-1.75	004.07	0
004	Fiber, Crude (%)	2188	3.110	0.0800	3.725	0.3718	0.1250	109	-1.65	004.07	0
004	Fiber, Crude (%)	0013	3.215	0.1700	3.725	0.3718	0.1250	109	-1.37	004.07	0
004	Fiber, Crude (%)	0175	3.280	0.0000	3.725	0.3718	0.1250	109	-1.20	004.00	0
004	Fiber, Crude (%)	0870	3.283	0.1697	3.725	0.3718	0.1250	109	-1.19	004.07	0
004	Fiber, Crude (%)	0589	3.305	0.0700	3.725	0.3718	0.1250	109	-1.13	004.07	0
004	Fiber, Crude (%)	0505	3.355	0.0900	3.725	0.3718	0.1250	109	-1.00	004.07	0
004	Fiber, Crude (%)	0644	3.362	0.0300	3.725	0.3718	0.1250	109	-0.98	004.07	0
004	Fiber, Crude (%)	0889	3.370	0.1100	3.725	0.3718	0.1250	109	-0.95	004.11	0
004	Fiber, Crude (%)	2181	3.380	0.0800	3.725	0.3718	0.1250	109	-0.93	004.07	0
004	Fiber, Crude (%)	2109	3.390	0.0200	3.725	0.3718	0.1250	109	-0.90	004.07	0
004	Fiber, Crude (%)	0098	3.395	0.3100	3.725	0.3718	0.1250	109	-0.89	004.07	0
004	Fiber, Crude (%)	0956	3.400	0.2000	3.725	0.3718	0.1250	109	-0.87	004.07	0
004	Fiber, Crude (%)	2066	3.405	0.0300	3.725	0.3718	0.1250	109	-0.86	004.00	0
004	Fiber, Crude (%)	2144	3.415	0.0300	3.725	0.3718	0.1250	109	-0.83	004.07	0
004	Fiber, Crude (%)	2053	3.430	0.1400	3.725	0.3718	0.1250	109	-0.79	004.07	0
004	Fiber, Crude (%)	0728	3.450	0.3000	3.725	0.3718	0.1250	109	-0.74	004.06	0
004	Fiber, Crude (%)	0190	3.450	0.1000	3.725	0.3718	0.1250	109	-0.74	004.07	0
004	Fiber, Crude (%)	0619	3.465	0.4100	3.725	0.3718	0.1250	109	-0.70	004.07	0
004	Fiber, Crude (%)	2304	3.470	0.0600	3.725	0.3718	0.1250	109	-0.69	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
004	Fiber, Crude (%)	2144	3.480	0.0400	3.725	0.3718	0.1250	109	-0.66	004.01	0
004	Fiber, Crude (%)	0004	3.490	0.2400	3.725	0.3718	0.1250	109	-0.63	004.07	0
004	Fiber, Crude (%)	0298	3.500	0.2000	3.725	0.3718	0.1250	109	-0.61	004.07	0
004	Fiber, Crude (%)	0918	3.510	0.0200	3.725	0.3718	0.1250	109	-0.58	004.06	0
004	Fiber, Crude (%)	0171	3.535	0.1700	3.725	0.3718	0.1250	109	-0.51	004.00	0
004	Fiber, Crude (%)	0009	3.540	0.1000	3.725	0.3718	0.1250	109	-0.50	004.07	0
004	Fiber, Crude (%)	0354	3.545	0.0100	3.725	0.3718	0.1250	109	-0.48	004.00	0
004	Fiber, Crude (%)	0689	3.550	0.1000	3.725	0.3718	0.1250	109	-0.47	004.06	0
004	Fiber, Crude (%)	0893	3.550	0.0000	3.725	0.3718	0.1250	109	-0.47	004.06	0
004	Fiber, Crude (%)	0708	3.550	0.0200	3.725	0.3718	0.1250	109	-0.47	004.07	0
004	Fiber, Crude (%)	2259	3.563	0.0600	3.725	0.3718	0.1250	109	-0.44	004.07	0
004	Fiber, Crude (%)	0921	3.570	0.5000	3.725	0.3718	0.1250	109	-0.42	004.03	0
004	Fiber, Crude (%)	0726	3.570	0.0000	3.725	0.3718	0.1250	109	-0.42	004.07	0
004	Fiber, Crude (%)	0083	3.575	0.1500	3.725	0.3718	0.1250	109	-0.40	004.07	0
004	Fiber, Crude (%)	0968	3.580	0.0200	3.725	0.3718	0.1250	109	-0.39	004.06	0
004	Fiber, Crude (%)	0098	3.585	0.1300	3.725	0.3718	0.1250	109	-0.38	004.06	0
004	Fiber, Crude (%)	0034	3.590	0.2800	3.725	0.3718	0.1250	109	-0.36	004.07	0
004	Fiber, Crude (%)	0169	3.600	0.2000	3.725	0.3718	0.1250	109	-0.34	004.00	0
004	Fiber, Crude (%)	0868	3.615	0.0300	3.725	0.3718	0.1250	109	-0.30	004.06	0
004	Fiber, Crude (%)	0848	3.620	0.0200	3.725	0.3718	0.1250	109	-0.28	004.07	0
004	Fiber, Crude (%)	0910	3.650	0.3000	3.725	0.3718	0.1250	109	-0.20	004.07	0
004	Fiber, Crude (%)	0164	3.650	0.1000	3.725	0.3718	0.1250	109	-0.20	004.00	0
004	Fiber, Crude (%)	0074	3.660	0.1000	3.725	0.3718	0.1250	109	-0.17	004.07	0
004	Fiber, Crude (%)	0353	3.665	0.0900	3.725	0.3718	0.1250	109	-0.16	004.03	0
004	Fiber, Crude (%)	0536	3.665	0.0900	3.725	0.3718	0.1250	109	-0.16	004.07	0
004	Fiber, Crude (%)	0881	3.667	0.1610	3.725	0.3718	0.1250	109	-0.16	004.06	0
004	Fiber, Crude (%)	0345	3.670	0.0400	3.725	0.3718	0.1250	109	-0.15	004.00	0
004	Fiber, Crude (%)	0100	3.695	0.2900	3.725	0.3718	0.1250	109	-0.08	004.07	0
004	Fiber, Crude (%)	0674	3.700	0.0000	3.725	0.3718	0.1250	109	-0.07	004.06	0
004	Fiber, Crude (%)	2022	3.700	0.0000	3.725	0.3718	0.1250	109	-0.07	004.06	0
004	Fiber, Crude (%)	0148	3.700	0.2000	3.725	0.3718	0.1250	109	-0.07	004.11	0
004	Fiber, Crude (%)	0553	3.700	0.1000	3.725	0.3718	0.1250	109	-0.07	004.11	0
004	Fiber, Crude (%)	0263	3.715	0.0400	3.725	0.3718	0.1250	109	-0.03	004.07	0
004	Fiber, Crude (%)	0675	3.720	0.1600	3.725	0.3718	0.1250	109	-0.01	004.07	0
004	Fiber, Crude (%)	0529	3.735	0.0700	3.725	0.3718	0.1250	109	0.03	004.07	0
004	Fiber, Crude (%)	0968	3.745	0.0700	3.725	0.3718	0.1250	109	0.05	004.07	0
004	Fiber, Crude (%)	0407	3.750	0.3678	3.725	0.3718	0.1250	109	0.07	004.07	0
004	Fiber, Crude (%)	2062	3.752	0.0080	3.725	0.3718	0.1250	109	0.07	004.06	0
004	Fiber, Crude (%)	2006	3.770	0.2800	3.725	0.3718	0.1250	109	0.12	004.06	0
004	Fiber, Crude (%)	0897	3.790	0.0200	3.725	0.3718	0.1250	109	0.18	004.06	0
004	Fiber, Crude (%)	2196	3.790	0.0000	3.725	0.3718	0.1250	109	0.18	004.07	0
004	Fiber, Crude (%)	0032	3.805	0.2300	3.725	0.3718	0.1250	109	0.22	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
004	Fiber, Crude (%)	0948	3.810	0.1000	3.725	0.3718	0.1250	109	0.23	004.07	0
004	Fiber, Crude (%)	0948	3.815	0.0100	3.725	0.3718	0.1250	109	0.24	004.06	0
004	Fiber, Crude (%)	2009	3.819	0.0681	3.725	0.3718	0.1250	109	0.25	004.07	0
004	Fiber, Crude (%)	0309	3.830	0.0050	3.725	0.3718	0.1250	109	0.28	004.00	0
004	Fiber, Crude (%)	0520	3.850	0.1000	3.725	0.3718	0.1250	109	0.34	004.07	0
004	Fiber, Crude (%)	0885	3.856	0.0179	3.725	0.3718	0.1250	109	0.35	004.03	0
004	Fiber, Crude (%)	0144	3.860	0.0600	3.725	0.3718	0.1250	109	0.36	004.07	0
004	Fiber, Crude (%)	0693	3.876	0.0770	3.725	0.3718	0.1250	109	0.40	004.07	0
004	Fiber, Crude (%)	0027	3.883	0.1590	3.725	0.3718	0.1250	109	0.42	004.06	0
004	Fiber, Crude (%)	0294	3.900	0.0000	3.725	0.3718	0.1250	109	0.47	004.07	0
004	Fiber, Crude (%)	0692	3.900	0.0000	3.725	0.3718	0.1250	109	0.47	004.07	0
004	Fiber, Crude (%)	2295	3.945	0.0100	3.725	0.3718	0.1250	109	0.59	004.06	0
004	Fiber, Crude (%)	0878	3.950	0.2200	3.725	0.3718	0.1250	109	0.61	004.06	0
004	Fiber, Crude (%)	0019	3.965	0.0500	3.725	0.3718	0.1250	109	0.65	004.07	0
004	Fiber, Crude (%)	0563	3.975	0.0457	3.725	0.3718	0.1250	109	0.67	004.07	0
004	Fiber, Crude (%)	0123	4.000	0.0600	3.725	0.3718	0.1250	109	0.74	004.07	0
004	Fiber, Crude (%)	0553	4.000	0.1600	3.725	0.3718	0.1250	109	0.74	004.07	0
004	Fiber, Crude (%)	0876	4.000	0.2000	3.725	0.3718	0.1250	109	0.74	004.07	0
004	Fiber, Crude (%)	2302	4.010	0.1400	3.725	0.3718	0.1250	109	0.77	004.07	0
004	Fiber, Crude (%)	2192	4.050	0.2200	3.725	0.3718	0.1250	109	0.87	004.03	0
004	Fiber, Crude (%)	0643	4.065	0.1100	3.725	0.3718	0.1250	109	0.91	004.07	0
004	Fiber, Crude (%)	2302	4.080	0.3000	3.725	0.3718	0.1250	109	0.96	004.11	0
004	Fiber, Crude (%)	2073	4.095	0.0300	3.725	0.3718	0.1250	109	1.00	004.00	0
004	Fiber, Crude (%)	2076	4.112	0.2990	3.725	0.3718	0.1250	109	1.04	004.00	0
004	Fiber, Crude (%)	0676	4.170	0.2600	3.725	0.3718	0.1250	109	1.20	004.06	0
004	Fiber, Crude (%)	0066	4.190	0.5000	3.725	0.3718	0.1250	109	1.25	004.07	0
004	Fiber, Crude (%)	0581	4.200	0.2000	3.725	0.3718	0.1250	109	1.28	004.07	0
004	Fiber, Crude (%)	2129	4.225	0.0521	3.725	0.3718	0.1250	109	1.35	004.99	0
004	Fiber, Crude (%)	0598	4.250	0.2200	3.725	0.3718	0.1250	109	1.41	004.07	0
004	Fiber, Crude (%)	0045	4.260	0.0200	3.725	0.3718	0.1250	109	1.44	004.07	0
004	Fiber, Crude (%)	0683	4.290	0.0400	3.725	0.3718	0.1250	109	1.52	004.07	0
004	Fiber, Crude (%)	0626	4.300	0.3200	3.725	0.3718	0.1250	109	1.55	004.07	0
004	Fiber, Crude (%)	0202	4.350	0.1200	3.725	0.3718	0.1250	109	1.68	004.07	0
004	Fiber, Crude (%)	0011	4.390	0.0000	3.725	0.3718	0.1250	109	1.79	004.11	0
004	Fiber, Crude (%)	0939	4.390	0.3600	3.725	0.3718	0.1250	109	1.79	004.07	0
004	Fiber, Crude (%)	2246	4.460	0.4200	3.725	0.3718	0.1250	109	1.98	004.03	0
004	Fiber, Crude (%)	0852	4.635	0.0700	3.725	0.3718	0.1250	109	2.45	004.11	0
004	Fiber, Crude (%)	0226	4.885	0.1900	3.725	0.3718	0.1250	109	3.12	004.00	0
004	Fiber, Crude (%)	0278	6.550	0.1000	3.725	0.3718	0.1250	109	7.60	004.07	0
004	Fiber, Crude (%)	0089	6.570	0.0000	3.725	0.3718	0.1250	109	7.65	004.07	0
004	Fiber, Crude (%)	0242	7.405	0.6100	3.725	0.3718	0.1250	109	9.90	004.07	0
004	Fiber, Crude (%)	0590	3.290	0.7800	3.725	0.3718	0.1250	109	-1.17	004.07	1

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
004	Fiber, Crude (%)	0028	3.350	0.9000	3.725	0.3718	0.1250	109	-1.01	004.07	1
005	Ash (%)	0148	6.100	0.0000	8.893	0.5216	0.0843	111	-5.35	005.11	0
005	Ash (%)	2081	7.040	0.0800	8.893	0.5216	0.0843	111	-3.55	005.99	0
005	Ash (%)	0853	7.450	0.3000	8.893	0.5216	0.0843	111	-2.77	005.00	0
005	Ash (%)	2184	7.685	0.3100	8.893	0.5216	0.0843	111	-2.32	005.00	0
005	Ash (%)	0144	7.860	0.0000	8.893	0.5216	0.0843	111	-1.98	005.00	0
005	Ash (%)	2268	7.895	0.0100	8.893	0.5216	0.0843	111	-1.91	005.00	0
005	Ash (%)	0169	7.915	0.0100	8.893	0.5216	0.0843	111	-1.87	005.00	0
005	Ash (%)	0852	7.925	0.0500	8.893	0.5216	0.0843	111	-1.86	005.11	0
005	Ash (%)	0539	7.945	0.0100	8.893	0.5216	0.0843	111	-1.82	005.00	0
005	Ash (%)	0278	7.950	0.0800	8.893	0.5216	0.0843	111	-1.81	005.00	0
005	Ash (%)	0357	8.000	0.0000	8.893	0.5216	0.0843	111	-1.71	005.00	0
005	Ash (%)	0425	8.030	0.0000	8.893	0.5216	0.0843	111	-1.65	005.00	0
005	Ash (%)	0541	8.065	0.2500	8.893	0.5216	0.0843	111	-1.59	005.00	0
005	Ash (%)	0918	8.160	0.0400	8.893	0.5216	0.0843	111	-1.40	005.00	0
005	Ash (%)	0042	8.175	0.0300	8.893	0.5216	0.0843	111	-1.38	005.00	0
005	Ash (%)	0062	8.179	0.2260	8.893	0.5216	0.0843	111	-1.37	005.00	0
005	Ash (%)	0019	8.200	0.4000	8.893	0.5216	0.0843	111	-1.33	005.00	0
005	Ash (%)	0229	8.215	0.4300	8.893	0.5216	0.0843	111	-1.30	005.00	0
005	Ash (%)	0598	8.240	0.0800	8.893	0.5216	0.0843	111	-1.25	005.00	0
005	Ash (%)	0953	8.240	0.0000	8.893	0.5216	0.0843	111	-1.25	005.00	0
005	Ash (%)	0574	8.255	0.0500	8.893	0.5216	0.0843	111	-1.22	005.00	0
005	Ash (%)	2302	8.265	0.0300	8.893	0.5216	0.0843	111	-1.20	005.00	0
005	Ash (%)	2073	8.340	0.1000	8.893	0.5216	0.0843	111	-1.06	005.00	0
005	Ash (%)	0164	8.360	0.0800	8.893	0.5216	0.0843	111	-1.02	005.00	0
005	Ash (%)	0175	8.365	0.0900	8.893	0.5216	0.0843	111	-1.01	005.00	0
005	Ash (%)	0726	8.425	0.1100	8.893	0.5216	0.0843	111	-0.90	005.05	0
005	Ash (%)	0956	8.450	0.1000	8.893	0.5216	0.0843	111	-0.85	005.00	0
005	Ash (%)	0662	8.549	0.0372	8.893	0.5216	0.0843	111	-0.66	005.99	0
005	Ash (%)	2022	8.590	0.0200	8.893	0.5216	0.0843	111	-0.58	005.00	0
005	Ash (%)	0263	8.599	0.0170	8.893	0.5216	0.0843	111	-0.56	005.05	0
005	Ash (%)	0682	8.620	0.0000	8.893	0.5216	0.0843	111	-0.52	005.00	0
005	Ash (%)	0589	8.625	0.4100	8.893	0.5216	0.0843	111	-0.51	005.00	0
005	Ash (%)	0358	8.660	0.0000	8.893	0.5216	0.0843	111	-0.45	005.00	0
005	Ash (%)	0186	8.684	0.0110	8.893	0.5216	0.0843	111	-0.40	005.00	0
005	Ash (%)	0297	8.705	0.2100	8.893	0.5216	0.0843	111	-0.36	005.11	0
005	Ash (%)	0893	8.750	0.0000	8.893	0.5216	0.0843	111	-0.27	005.05	0
005	Ash (%)	0142	8.755	0.0700	8.893	0.5216	0.0843	111	-0.26	005.00	0
005	Ash (%)	0083	8.760	0.0200	8.893	0.5216	0.0843	111	-0.25	005.00	0
005	Ash (%)	0870	8.768	0.0482	8.893	0.5216	0.0843	111	-0.24	005.00	0
005	Ash (%)	0536	8.820	0.0200	8.893	0.5216	0.0843	111	-0.14	005.99	0
005	Ash (%)	0035	8.845	0.0100	8.893	0.5216	0.0843	111	-0.09	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
005	Ash (%)	0563	8.849	0.0321	8.893	0.5216	0.0843	111	-0.08	005.00	0
005	Ash (%)	0619	8.880	0.0400	8.893	0.5216	0.0843	111	-0.02	005.05	0
005	Ash (%)	0656	8.880	0.0000	8.893	0.5216	0.0843	111	-0.02	005.00	0
005	Ash (%)	2150	8.885	0.0100	8.893	0.5216	0.0843	111	-0.01	005.00	0
005	Ash (%)	0098	8.915	0.1500	8.893	0.5216	0.0843	111	0.04	005.00	0
005	Ash (%)	0123	8.915	0.1500	8.893	0.5216	0.0843	111	0.04	005.00	0
005	Ash (%)	0354	8.955	0.0100	8.893	0.5216	0.0843	111	0.12	005.00	0
005	Ash (%)	0968	8.970	0.0200	8.893	0.5216	0.0843	111	0.15	005.00	0
005	Ash (%)	2196	8.980	0.0000	8.893	0.5216	0.0843	111	0.17	005.00	0
005	Ash (%)	0546	8.985	0.0100	8.893	0.5216	0.0843	111	0.18	005.99	0
005	Ash (%)	0108	9.000	0.2400	8.893	0.5216	0.0843	111	0.21	005.00	0
005	Ash (%)	0298	9.000	0.0200	8.893	0.5216	0.0843	111	0.21	005.00	0
005	Ash (%)	0510	9.010	0.1400	8.893	0.5216	0.0843	111	0.23	005.00	0
005	Ash (%)	0618	9.011	0.2670	8.893	0.5216	0.0843	111	0.23	005.00	0
005	Ash (%)	0693	9.036	0.0680	8.893	0.5216	0.0843	111	0.28	005.00	0
005	Ash (%)	0265	9.050	0.1000	8.893	0.5216	0.0843	111	0.30	005.00	0
005	Ash (%)	0881	9.064	0.1600	8.893	0.5216	0.0843	111	0.33	005.00	0
005	Ash (%)	0968	9.070	0.0200	8.893	0.5216	0.0843	111	0.34	005.02	0
005	Ash (%)	2109	9.070	0.0000	8.893	0.5216	0.0843	111	0.34	005.05	0
005	Ash (%)	2192	9.080	0.3800	8.893	0.5216	0.0843	111	0.36	005.05	0
005	Ash (%)	0897	9.090	0.0200	8.893	0.5216	0.0843	111	0.38	005.05	0
005	Ash (%)	2006	9.095	0.0100	8.893	0.5216	0.0843	111	0.39	005.00	0
005	Ash (%)	0407	9.099	0.3072	8.893	0.5216	0.0843	111	0.40	005.00	0
005	Ash (%)	0683	9.100	0.1200	8.893	0.5216	0.0843	111	0.40	005.05	0
005	Ash (%)	0889	9.110	0.1000	8.893	0.5216	0.0843	111	0.42	005.05	0
005	Ash (%)	0353	9.115	0.0700	8.893	0.5216	0.0843	111	0.43	005.00	0
005	Ash (%)	0242	9.130	0.0400	8.893	0.5216	0.0843	111	0.46	005.00	0
005	Ash (%)	2076	9.133	0.0005	8.893	0.5216	0.0843	111	0.46	005.00	0
005	Ash (%)	0100	9.140	0.2200	8.893	0.5216	0.0843	111	0.47	005.00	0
005	Ash (%)	2188	9.145	0.0100	8.893	0.5216	0.0843	111	0.48	005.00	0
005	Ash (%)	2012	9.145	0.0300	8.893	0.5216	0.0843	111	0.48	005.05	0
005	Ash (%)	0297	9.150	0.1000	8.893	0.5216	0.0843	111	0.49	005.00	0
005	Ash (%)	0689	9.150	0.0200	8.893	0.5216	0.0843	111	0.49	005.05	0
005	Ash (%)	0148	9.160	0.1400	8.893	0.5216	0.0843	111	0.51	005.00	0
005	Ash (%)	0034	9.164	0.1420	8.893	0.5216	0.0843	111	0.52	005.05	0
005	Ash (%)	2181	9.195	0.0500	8.893	0.5216	0.0843	111	0.58	005.00	0
005	Ash (%)	0692	9.200	0.0000	8.893	0.5216	0.0843	111	0.59	005.05	0
005	Ash (%)	2259	9.203	0.0200	8.893	0.5216	0.0843	111	0.60	005.00	0
005	Ash (%)	0675	9.215	0.0500	8.893	0.5216	0.0843	111	0.62	005.00	0
005	Ash (%)	0885	9.220	0.0972	8.893	0.5216	0.0843	111	0.63	005.05	0
005	Ash (%)	0948	9.225	0.0100	8.893	0.5216	0.0843	111	0.64	005.00	0
005	Ash (%)	0171	9.230	0.0400	8.893	0.5216	0.0843	111	0.65	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
005	Ash (%)	0643	9.230	0.1400	8.893	0.5216	0.0843	111	0.65	005.00	0
005	Ash (%)	2295	9.235	0.0100	8.893	0.5216	0.0843	111	0.66	005.05	0
005	Ash (%)	0674	9.240	0.0000	8.893	0.5216	0.0843	111	0.67	005.00	0
005	Ash (%)	0309	9.267	0.0280	8.893	0.5216	0.0843	111	0.72	005.00	0
005	Ash (%)	2291	9.290	0.3400	8.893	0.5216	0.0843	111	0.76	005.05	0
005	Ash (%)	0644	9.294	0.0240	8.893	0.5216	0.0843	111	0.77	005.02	0
005	Ash (%)	0868	9.305	0.0300	8.893	0.5216	0.0843	111	0.79	005.05	0
005	Ash (%)	0676	9.325	0.0300	8.893	0.5216	0.0843	111	0.83	005.99	0
005	Ash (%)	2062	9.334	0.0361	8.893	0.5216	0.0843	111	0.85	005.05	0
005	Ash (%)	0202	9.350	0.0200	8.893	0.5216	0.0843	111	0.88	005.99	0
005	Ash (%)	0910	9.350	0.3000	8.893	0.5216	0.0843	111	0.88	005.99	0
005	Ash (%)	0581	9.350	0.1000	8.893	0.5216	0.0843	111	0.88	005.00	0
005	Ash (%)	2146	9.360	0.0800	8.893	0.5216	0.0843	111	0.90	005.05	0
005	Ash (%)	0045	9.365	0.1700	8.893	0.5216	0.0843	111	0.91	005.00	0
005	Ash (%)	0878	9.365	0.0300	8.893	0.5216	0.0843	111	0.91	005.05	0
005	Ash (%)	0294	9.400	0.0000	8.893	0.5216	0.0843	111	0.97	005.05	0
005	Ash (%)	0226	9.405	0.2100	8.893	0.5216	0.0843	111	0.98	005.00	0
005	Ash (%)	0520	9.415	0.0100	8.893	0.5216	0.0843	111	1.00	005.00	0
005	Ash (%)	0590	9.415	0.1100	8.893	0.5216	0.0843	111	1.00	005.05	0
005	Ash (%)	0728	9.420	0.0800	8.893	0.5216	0.0843	111	1.01	005.99	0
005	Ash (%)	0652	9.450	0.1000	8.893	0.5216	0.0843	111	1.07	005.99	0
005	Ash (%)	2009	9.452	0.0080	8.893	0.5216	0.0843	111	1.07	005.05	0
005	Ash (%)	0345	9.460	0.0800	8.893	0.5216	0.0843	111	1.09	005.00	0
005	Ash (%)	2171	9.465	0.0700	8.893	0.5216	0.0843	111	1.10	005.00	0
005	Ash (%)	2246	9.500	0.0200	8.893	0.5216	0.0843	111	1.16	005.05	0
005	Ash (%)	2129	9.634	0.0526	8.893	0.5216	0.0843	111	1.42	005.05	0
005	Ash (%)	0852	9.670	0.0200	8.893	0.5216	0.0843	111	1.49	005.00	0
005	Ash (%)	0889	11.02	0.1000	8.893	0.5216	0.0843	111	4.08	005.11	0
005	Ash (%)	0553	8.520	1.000	8.893	0.5216	0.0843	111	-0.71	005.00	1
006	Total Sugars (%)	0407	3.554	0.0430	3.728	0.1594	0.0858	4		006.00	0
006	Total Sugars (%)	0910	3.640	0.0200	3.728	0.1594	0.0858	4		006.00	0
006	Total Sugars (%)	0941	3.820	0.0800	3.728	0.1594	0.0858	4		006.03	0
006	Total Sugars (%)	0956	3.900	0.2000	3.728	0.1594	0.0858	4		006.99	0
006	Total Sugars (%)	2216	3.490	1.040	3.728	0.1594	0.0858	4		006.99	1
008	Fiber, Acid Detergent (%)	0956	4.250	0.5000	5.052	0.3852	0.1823	51	-2.08	008.08	0
008	Fiber, Acid Detergent (%)	2012	4.265	0.0700	5.052	0.3852	0.1823	51	-2.04	008.08	0
008	Fiber, Acid Detergent (%)	2302	4.460	0.3800	5.052	0.3852	0.1823	51	-1.54	008.08	0
008	Fiber, Acid Detergent (%)	0590	4.590	0.4000	5.052	0.3852	0.1823	51	-1.20	008.08	0
008	Fiber, Acid Detergent (%)	0353	4.615	0.1900	5.052	0.3852	0.1823	51	-1.13	008.02	0
008	Fiber, Acid Detergent (%)	0848	4.625	0.0300	5.052	0.3852	0.1823	51	-1.11	008.08	0
008	Fiber, Acid Detergent (%)	0948	4.655	0.1500	5.052	0.3852	0.1823	51	-1.03	008.02	0
008	Fiber, Acid Detergent (%)	0726	4.705	0.1500	5.052	0.3852	0.1823	51	-0.90	008.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
008	Fiber, Acid Detergent (%)	0098	4.725	0.2500	5.052	0.3852	0.1823	51	-0.85	008.02	0
008	Fiber, Acid Detergent (%)	0968	4.735	0.1700	5.052	0.3852	0.1823	51	-0.82	008.08	0
008	Fiber, Acid Detergent (%)	0910	4.740	0.2000	5.052	0.3852	0.1823	51	-0.81	008.08	0
008	Fiber, Acid Detergent (%)	0278	4.750	0.1000	5.052	0.3852	0.1823	51	-0.78	008.08	0
008	Fiber, Acid Detergent (%)	0948	4.770	0.2400	5.052	0.3852	0.1823	51	-0.73	008.08	0
008	Fiber, Acid Detergent (%)	0190	4.840	0.2400	5.052	0.3852	0.1823	51	-0.55	008.08	0
008	Fiber, Acid Detergent (%)	0226	4.845	0.1100	5.052	0.3852	0.1823	51	-0.54	008.02	0
008	Fiber, Acid Detergent (%)	0345	4.850	0.1000	5.052	0.3852	0.1823	51	-0.52	008.02	0
008	Fiber, Acid Detergent (%)	0294	4.850	0.1000	5.052	0.3852	0.1823	51	-0.52	008.08	0
008	Fiber, Acid Detergent (%)	0027	4.857	0.0090	5.052	0.3852	0.1823	51	-0.51	008.99	0
008	Fiber, Acid Detergent (%)	0309	4.862	0.0010	5.052	0.3852	0.1823	51	-0.49	008.02	0
008	Fiber, Acid Detergent (%)	0083	4.865	0.0500	5.052	0.3852	0.1823	51	-0.49	008.08	0
008	Fiber, Acid Detergent (%)	0619	4.875	0.1900	5.052	0.3852	0.1823	51	-0.46	008.08	0
008	Fiber, Acid Detergent (%)	0004	4.890	0.0800	5.052	0.3852	0.1823	51	-0.42	008.08	0
008	Fiber, Acid Detergent (%)	0357	4.900	0.4000	5.052	0.3852	0.1823	51	-0.39	008.08	0
008	Fiber, Acid Detergent (%)	0675	4.925	0.2900	5.052	0.3852	0.1823	51	-0.33	008.08	0
008	Fiber, Acid Detergent (%)	0037	5.025	0.0900	5.052	0.3852	0.1823	51	-0.07	008.08	0
008	Fiber, Acid Detergent (%)	0728	5.040	0.2400	5.052	0.3852	0.1823	51	-0.03	008.02	0
008	Fiber, Acid Detergent (%)	0358	5.060	0.4600	5.052	0.3852	0.1823	51	0.02	008.08	0
008	Fiber, Acid Detergent (%)	2196	5.090	0.0000	5.052	0.3852	0.1823	51	0.10	008.08	0
008	Fiber, Acid Detergent (%)	0689	5.150	0.1000	5.052	0.3852	0.1823	51	0.25	008.02	0
008	Fiber, Acid Detergent (%)	0164	5.150	0.7000	5.052	0.3852	0.1823	51	0.25	008.08	0
008	Fiber, Acid Detergent (%)	0066	5.160	0.1800	5.052	0.3852	0.1823	51	0.28	008.08	0
008	Fiber, Acid Detergent (%)	0148	5.160	0.0600	5.052	0.3852	0.1823	51	0.28	008.08	0
008	Fiber, Acid Detergent (%)	0510	5.200	0.2000	5.052	0.3852	0.1823	51	0.38	008.08	0
008	Fiber, Acid Detergent (%)	0001	5.210	0.3800	5.052	0.3852	0.1823	51	0.41	008.08	0
008	Fiber, Acid Detergent (%)	0581	5.215	0.1500	5.052	0.3852	0.1823	51	0.42	008.08	0
008	Fiber, Acid Detergent (%)	0889	5.220	0.0400	5.052	0.3852	0.1823	51	0.44	008.08	0
008	Fiber, Acid Detergent (%)	0297	5.230	0.0200	5.052	0.3852	0.1823	51	0.46	008.08	0
008	Fiber, Acid Detergent (%)	0878	5.300	0.3400	5.052	0.3852	0.1823	51	0.64	008.08	0
008	Fiber, Acid Detergent (%)	0148	5.350	0.2000	5.052	0.3852	0.1823	51	0.77	008.02	0
008	Fiber, Acid Detergent (%)	0263	5.372	0.0240	5.052	0.3852	0.1823	51	0.83	008.08	0
008	Fiber, Acid Detergent (%)	0407	5.378	0.1210	5.052	0.3852	0.1823	51	0.85	008.08	0
008	Fiber, Acid Detergent (%)	0536	5.385	0.0900	5.052	0.3852	0.1823	51	0.86	008.08	0
008	Fiber, Acid Detergent (%)	2246	5.420	0.1200	5.052	0.3852	0.1823	51	0.96	008.02	0
008	Fiber, Acid Detergent (%)	0354	5.420	0.0200	5.052	0.3852	0.1823	51	0.96	008.08	0
008	Fiber, Acid Detergent (%)	2129	5.554	0.0985	5.052	0.3852	0.1823	51	1.30	008.08	0
008	Fiber, Acid Detergent (%)	0870	5.566	0.3828	5.052	0.3852	0.1823	51	1.33	008.08	0
008	Fiber, Acid Detergent (%)	0693	5.582	0.2000	5.052	0.3852	0.1823	51	1.38	008.08	0
008	Fiber, Acid Detergent (%)	0045	5.680	0.0800	5.052	0.3852	0.1823	51	1.63	008.08	0
008	Fiber, Acid Detergent (%)	0964	5.931	0.1000	5.052	0.3852	0.1823	51	2.28	008.02	0
008	Fiber, Acid Detergent (%)	0918	6.031	0.1125	5.052	0.3852	0.1823	51	2.54	008.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
008	Fiber, Acid Detergent (%)	0265	6.225	0.3900	5.052	0.3852	0.1823	51	3.05	008.05	0
008	Fiber, Acid Detergent (%)	0013	4.630	0.8400	5.052	0.3852	0.1823	51	-1.10	008.08	1
009	Fiber, Neutral Detergent (%)	0910	10.80	0.0000	12.80	0.9173	0.3396	50	-2.18	009.07	0
009	Fiber, Neutral Detergent (%)	0598	11.05	0.0400	12.80	0.9173	0.3396	50	-1.91	009.09	0
009	Fiber, Neutral Detergent (%)	2129	11.06	1.481	12.80	0.9173	0.3396	50	-1.90	009.09	0
009	Fiber, Neutral Detergent (%)	0939	11.37	0.0300	12.80	0.9173	0.3396	50	-1.57	009.09	0
009	Fiber, Neutral Detergent (%)	0263	11.75	0.0180	12.80	0.9173	0.3396	50	-1.15	009.09	0
009	Fiber, Neutral Detergent (%)	0956	11.75	0.7000	12.80	0.9173	0.3396	50	-1.15	009.09	0
009	Fiber, Neutral Detergent (%)	0358	11.78	0.3100	12.80	0.9173	0.3396	50	-1.12	009.09	0
009	Fiber, Neutral Detergent (%)	0675	12.16	0.0700	12.80	0.9173	0.3396	50	-0.71	009.09	0
009	Fiber, Neutral Detergent (%)	0870	12.23	0.2234	12.80	0.9173	0.3396	50	-0.63	009.09	0
009	Fiber, Neutral Detergent (%)	0948	12.23	0.0600	12.80	0.9173	0.3396	50	-0.62	009.07	0
009	Fiber, Neutral Detergent (%)	0083	12.25	0.0600	12.80	0.9173	0.3396	50	-0.60	009.09	0
009	Fiber, Neutral Detergent (%)	0878	12.26	0.0600	12.80	0.9173	0.3396	50	-0.59	009.09	0
009	Fiber, Neutral Detergent (%)	0353	12.30	0.2500	12.80	0.9173	0.3396	50	-0.55	009.07	0
009	Fiber, Neutral Detergent (%)	2302	12.30	0.6800	12.80	0.9173	0.3396	50	-0.55	009.09	0
009	Fiber, Neutral Detergent (%)	0948	12.32	0.0500	12.80	0.9173	0.3396	50	-0.53	009.09	0
009	Fiber, Neutral Detergent (%)	0407	12.33	0.1940	12.80	0.9173	0.3396	50	-0.52	009.09	0
009	Fiber, Neutral Detergent (%)	0098	12.35	0.5000	12.80	0.9173	0.3396	50	-0.49	009.07	0
009	Fiber, Neutral Detergent (%)	2012	12.38	0.5500	12.80	0.9173	0.3396	50	-0.47	009.09	0
009	Fiber, Neutral Detergent (%)	0278	12.40	0.2000	12.80	0.9173	0.3396	50	-0.44	009.09	0
009	Fiber, Neutral Detergent (%)	0918	12.42	0.2075	12.80	0.9173	0.3396	50	-0.41	009.09	0
009	Fiber, Neutral Detergent (%)	0581	12.44	0.2100	12.80	0.9173	0.3396	50	-0.40	009.09	0
009	Fiber, Neutral Detergent (%)	0294	12.45	0.1000	12.80	0.9173	0.3396	50	-0.38	009.09	0
009	Fiber, Neutral Detergent (%)	0357	12.45	0.5000	12.80	0.9173	0.3396	50	-0.38	009.09	0
009	Fiber, Neutral Detergent (%)	0889	12.57	0.2500	12.80	0.9173	0.3396	50	-0.26	009.99	0
009	Fiber, Neutral Detergent (%)	0726	12.58	0.1300	12.80	0.9173	0.3396	50	-0.25	009.09	0
009	Fiber, Neutral Detergent (%)	0693	12.73	0.6740	12.80	0.9173	0.3396	50	-0.08	009.09	0
009	Fiber, Neutral Detergent (%)	0037	12.78	0.0000	12.80	0.9173	0.3396	50	-0.03	009.09	0
009	Fiber, Neutral Detergent (%)	0148	12.86	0.1400	12.80	0.9173	0.3396	50	0.06	009.09	0
009	Fiber, Neutral Detergent (%)	0510	12.90	0.6000	12.80	0.9173	0.3396	50	0.11	009.09	0
009	Fiber, Neutral Detergent (%)	0066	12.91	0.2100	12.80	0.9173	0.3396	50	0.11	009.09	0
009	Fiber, Neutral Detergent (%)	0345	12.95	0.3000	12.80	0.9173	0.3396	50	0.16	009.07	0
009	Fiber, Neutral Detergent (%)	0226	12.96	0.1900	12.80	0.9173	0.3396	50	0.17	009.07	0
009	Fiber, Neutral Detergent (%)	0619	13.12	0.3300	12.80	0.9173	0.3396	50	0.34	009.09	0
009	Fiber, Neutral Detergent (%)	0297	13.15	0.5400	12.80	0.9173	0.3396	50	0.38	009.09	0
009	Fiber, Neutral Detergent (%)	0265	13.25	0.1000	12.80	0.9173	0.3396	50	0.49	009.09	0
009	Fiber, Neutral Detergent (%)	0354	13.38	0.0100	12.80	0.9173	0.3396	50	0.62	009.09	0
009	Fiber, Neutral Detergent (%)	0536	13.42	0.0800	12.80	0.9173	0.3396	50	0.67	009.09	0
009	Fiber, Neutral Detergent (%)	0045	13.45	0.9000	12.80	0.9173	0.3396	50	0.71	009.09	0
009	Fiber, Neutral Detergent (%)	0164	13.45	1.900	12.80	0.9173	0.3396	50	0.71	009.09	0
009	Fiber, Neutral Detergent (%)	0881	13.45	0.1440	12.80	0.9173	0.3396	50	0.71	009.09	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
009	Fiber, Neutral Detergent (%)	2196	13.62	0.0000	12.80	0.9173	0.3396	50	0.89	009.09	0
009	Fiber, Neutral Detergent (%)	0590	13.75	1.790	12.80	0.9173	0.3396	50	1.03	009.09	0
009	Fiber, Neutral Detergent (%)	0848	13.81	0.1900	12.80	0.9173	0.3396	50	1.09	009.09	0
009	Fiber, Neutral Detergent (%)	0941	13.81	0.0000	12.80	0.9173	0.3396	50	1.10	009.99	0
009	Fiber, Neutral Detergent (%)	0968	14.17	0.6000	12.80	0.9173	0.3396	50	1.49	009.09	0
009	Fiber, Neutral Detergent (%)	0309	14.66	0.0798	12.80	0.9173	0.3396	50	2.03	009.07	0
009	Fiber, Neutral Detergent (%)	0689	14.70	0.4000	12.80	0.9173	0.3396	50	2.07	009.07	0
009	Fiber, Neutral Detergent (%)	2246	15.49	0.3700	12.80	0.9173	0.3396	50	2.92	009.07	0
009	Fiber, Neutral Detergent (%)	0190	16.61	0.5400	12.80	0.9173	0.3396	50	4.15	009.09	0
009	Fiber, Neutral Detergent (%)	0889	16.80	0.0200	12.80	0.9173	0.3396	50	4.36	009.09	0
009	Fiber, Neutral Detergent (%)	0964	11.47	3.081	12.80	0.9173	0.3396	50	-1.46	009.07	1
009	Fiber, Neutral Detergent (%)	2246	37.07	0.8100	12.80	0.9173	0.3396	50	26.45	009.04	2
010	Moisture (%)	0692	8.900	0.2000	9.584	0.4879	0.0878	19	-1.40	010.99	0
010	Moisture (%)	0546	8.960	0.0200	9.584	0.4879	0.0878	19	-1.28	010.99	0
010	Moisture (%)	0726	9.090	0.1800	9.584	0.4879	0.0878	19	-1.01	010.99	0
010	Moisture (%)	0948	9.165	0.0500	9.584	0.4879	0.0878	19	-0.86	010.99	0
010	Moisture (%)	0939	9.240	0.1000	9.584	0.4879	0.0878	19	-0.70	010.99	0
010	Moisture (%)	0889	9.335	0.0300	9.584	0.4879	0.0878	19	-0.51	010.11	0
010	Moisture (%)	2129	9.350	0.0385	9.584	0.4879	0.0878	19	-0.48	010.99	0
010	Moisture (%)	2043	9.450	0.0800	9.584	0.4879	0.0878	19	-0.27	010.99	0
010	Moisture (%)	2076	9.555	0.0700	9.584	0.4879	0.0878	19	-0.06	010.99	0
010	Moisture (%)	0921	9.580	0.1200	9.584	0.4879	0.0878	19	-0.01	010.99	0
010	Moisture (%)	0652	9.600	0.0000	9.584	0.4879	0.0878	19	0.03	010.99	0
010	Moisture (%)	2178	9.625	0.0100	9.584	0.4879	0.0878	19	0.08	010.99	0
010	Moisture (%)	0914	9.720	0.1400	9.584	0.4879	0.0878	19	0.28	010.99	0
010	Moisture (%)	2150	9.745	0.0300	9.584	0.4879	0.0878	19	0.33	010.99	0
010	Moisture (%)	0190	9.955	0.0100	9.584	0.4879	0.0878	19	0.76	010.99	0
010	Moisture (%)	0164	9.985	0.0500	9.584	0.4879	0.0878	19	0.82	010.03	0
010	Moisture (%)	0148	10.21	0.0900	9.584	0.4879	0.0878	19	1.27	010.11	0
010	Moisture (%)	2302	10.48	0.4400	9.584	0.4879	0.0878	19	1.84	010.11	0
010	Moisture (%)	0852	10.89	0.0100	9.584	0.4879	0.0878	19	2.67	010.11	0
010	Moisture (%)	0843	10.21	1.530	9.584	0.4879	0.0878	19	1.27	010.03	1
011	Loss on Drying (%)	0520	8.850	0.1000	10.43	0.5075	0.1045	63	-3.11	011.01	0
011	Loss on Drying (%)	0536	9.525	0.0900	10.43	0.5075	0.1045	63	-1.78	011.01	0
011	Loss on Drying (%)	0425	9.570	0.0000	10.43	0.5075	0.1045	63	-1.69	011.01	0
011	Loss on Drying (%)	0175	9.650	0.1000	10.43	0.5075	0.1045	63	-1.53	011.01	0
011	Loss on Drying (%)	0968	9.710	0.0200	10.43	0.5075	0.1045	63	-1.41	011.01	0
011	Loss on Drying (%)	2066	9.755	0.0100	10.43	0.5075	0.1045	63	-1.32	011.02	0
011	Loss on Drying (%)	0953	9.905	0.0500	10.43	0.5075	0.1045	63	-1.03	011.01	0
011	Loss on Drying (%)	0859	9.960	0.0390	10.43	0.5075	0.1045	63	-0.92	011.01	0
011	Loss on Drying (%)	2303	9.962	0.1140	10.43	0.5075	0.1045	63	-0.92	011.99	0
011	Loss on Drying (%)	0881	9.969	0.0800	10.43	0.5075	0.1045	63	-0.90	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
011	Loss on Drying (%)	0294	10.00	0.0000	10.43	0.5075	0.1045	63	-0.84	011.01	0
011	Loss on Drying (%)	0675	10.02	0.0100	10.43	0.5075	0.1045	63	-0.81	011.01	0
011	Loss on Drying (%)	0354	10.02	0.0400	10.43	0.5075	0.1045	63	-0.80	011.01	0
011	Loss on Drying (%)	0598	10.03	0.0100	10.43	0.5075	0.1045	63	-0.79	011.01	0
011	Loss on Drying (%)	0298	10.04	0.0200	10.43	0.5075	0.1045	63	-0.76	011.01	0
011	Loss on Drying (%)	2188	10.04	0.0000	10.43	0.5075	0.1045	63	-0.76	011.01	0
011	Loss on Drying (%)	0857	10.05	0.1000	10.43	0.5075	0.1045	63	-0.74	011.99	0
011	Loss on Drying (%)	2168	10.06	0.2500	10.43	0.5075	0.1045	63	-0.73	011.01	0
011	Loss on Drying (%)	2174	10.08	0.0300	10.43	0.5075	0.1045	63	-0.69	011.01	0
011	Loss on Drying (%)	0563	10.10	0.0044	10.43	0.5075	0.1045	63	-0.65	011.01	0
011	Loss on Drying (%)	0100	10.12	0.0400	10.43	0.5075	0.1045	63	-0.60	011.01	0
011	Loss on Drying (%)	0263	10.13	0.0300	10.43	0.5075	0.1045	63	-0.59	011.01	0
011	Loss on Drying (%)	0062	10.16	0.0800	10.43	0.5075	0.1045	63	-0.53	011.01	0
011	Loss on Drying (%)	0942	10.18	0.0800	10.43	0.5075	0.1045	63	-0.49	011.02	0
011	Loss on Drying (%)	0098	10.22	0.1200	10.43	0.5075	0.1045	63	-0.41	011.01	0
011	Loss on Drying (%)	0539	10.24	0.0200	10.43	0.5075	0.1045	63	-0.37	011.01	0
011	Loss on Drying (%)	2006	10.25	0.1200	10.43	0.5075	0.1045	63	-0.35	011.01	0
011	Loss on Drying (%)	0674	10.31	0.0000	10.43	0.5075	0.1045	63	-0.23	011.01	0
011	Loss on Drying (%)	2196	10.33	0.0000	10.43	0.5075	0.1045	63	-0.19	011.01	0
011	Loss on Drying (%)	0164	10.42	0.5600	10.43	0.5075	0.1045	63	-0.01	011.01	0
011	Loss on Drying (%)	0574	10.43	0.0400	10.43	0.5075	0.1045	63	0.01	011.01	0
011	Loss on Drying (%)	0510	10.45	0.1000	10.43	0.5075	0.1045	63	0.05	011.01	0
011	Loss on Drying (%)	0728	10.46	0.0000	10.43	0.5075	0.1045	63	0.07	011.01	0
011	Loss on Drying (%)	0226	10.49	0.0500	10.43	0.5075	0.1045	63	0.11	011.01	0
011	Loss on Drying (%)	2291	10.51	0.1300	10.43	0.5075	0.1045	63	0.15	011.01	0
011	Loss on Drying (%)	0066	10.55	0.1600	10.43	0.5075	0.1045	63	0.24	011.01	0
011	Loss on Drying (%)	0553	10.56	0.1500	10.43	0.5075	0.1045	63	0.25	011.01	0
011	Loss on Drying (%)	0852	10.56	0.0400	10.43	0.5075	0.1045	63	0.26	011.99	0
011	Loss on Drying (%)	0870	10.57	0.2473	10.43	0.5075	0.1045	63	0.28	011.01	0
011	Loss on Drying (%)	2181	10.59	0.1200	10.43	0.5075	0.1045	63	0.32	011.01	0
011	Loss on Drying (%)	0643	10.60	0.0000	10.43	0.5075	0.1045	63	0.34	011.01	0
011	Loss on Drying (%)	0171	10.62	0.1500	10.43	0.5075	0.1045	63	0.37	011.01	0
011	Loss on Drying (%)	0529	10.67	0.1000	10.43	0.5075	0.1045	63	0.48	011.02	0
011	Loss on Drying (%)	2073	10.75	0.0900	10.43	0.5075	0.1045	63	0.63	011.01	0
011	Loss on Drying (%)	2022	10.75	0.0300	10.43	0.5075	0.1045	63	0.63	011.01	0
011	Loss on Drying (%)	0229	10.81	0.1800	10.43	0.5075	0.1045	63	0.75	011.01	0
011	Loss on Drying (%)	0618	10.81	0.0000	10.43	0.5075	0.1045	63	0.75	011.01	0
011	Loss on Drying (%)	2302	10.87	0.0700	10.43	0.5075	0.1045	63	0.86	011.01	0
011	Loss on Drying (%)	0309	10.87	0.1400	10.43	0.5075	0.1045	63	0.87	011.01	0
011	Loss on Drying (%)	0242	10.89	0.1200	10.43	0.5075	0.1045	63	0.91	011.01	0
011	Loss on Drying (%)	0682	10.89	0.0000	10.43	0.5075	0.1045	63	0.91	011.01	0
011	Loss on Drying (%)	0144	10.91	0.1100	10.43	0.5075	0.1045	63	0.94	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
011	Loss on Drying (%)	0123	10.93	0.0400	10.43	0.5075	0.1045	63	0.99	011.01	0
011	Loss on Drying (%)	0541	10.94	0.0600	10.43	0.5075	0.1045	63	1.01	011.01	0
011	Loss on Drying (%)	2268	10.95	0.0100	10.43	0.5075	0.1045	63	1.02	011.01	0
011	Loss on Drying (%)	0265	10.95	0.7000	10.43	0.5075	0.1045	63	1.03	011.01	0
011	Loss on Drying (%)	0108	10.98	0.0400	10.43	0.5075	0.1045	63	1.09	011.01	0
011	Loss on Drying (%)	0589	11.05	0.0700	10.43	0.5075	0.1045	63	1.22	011.01	0
011	Loss on Drying (%)	0004	11.15	0.0400	10.43	0.5075	0.1045	63	1.42	011.01	0
011	Loss on Drying (%)	0358	11.16	0.6300	10.43	0.5075	0.1045	63	1.43	011.01	0
011	Loss on Drying (%)	0202	11.17	0.0900	10.43	0.5075	0.1045	63	1.45	011.01	0
011	Loss on Drying (%)	2184	11.23	0.1600	10.43	0.5075	0.1045	63	1.58	011.01	0
011	Loss on Drying (%)	0843	11.67	0.6000	10.43	0.5075	0.1045	63	2.45	011.01	0
011	Loss on Drying (%)	0407	10.52	0.8452	10.43	0.5075	0.1045	63	0.18	011.01	1
011	Loss on Drying (%)	0013	10.63	0.7100	10.43	0.5075	0.1045	63	0.39	011.01	1
012	Starch (%)	0164	35.95	0.7000	39.98	2.293	0.5604	34	-1.76	012.01	0
012	Starch (%)	0889	36.19	0.1100	39.98	2.293	0.5604	34	-1.66	012.11	0
012	Starch (%)	0918	36.20	2.412	39.98	2.293	0.5604	34	-1.65	012.01	0
012	Starch (%)	0098	36.45	1.900	39.98	2.293	0.5604	34	-1.54	012.04	0
012	Starch (%)	0910	36.80	0.0000	39.98	2.293	0.5604	34	-1.39	012.01	0
012	Starch (%)	0956	37.00	1.000	39.98	2.293	0.5604	34	-1.30	012.01	0
012	Starch (%)	0148	37.80	0.6600	39.98	2.293	0.5604	34	-0.95	012.01	0
012	Starch (%)	0510	37.85	0.1000	39.98	2.293	0.5604	34	-0.93	012.04	0
012	Starch (%)	0870	38.01	0.7265	39.98	2.293	0.5604	34	-0.86	012.01	0
012	Starch (%)	0045	39.45	0.9000	39.98	2.293	0.5604	34	-0.23	012.01	0
012	Starch (%)	0297	39.80	2.190	39.98	2.293	0.5604	34	-0.08	012.03	0
012	Starch (%)	0852	39.93	0.1500	39.98	2.293	0.5604	34	-0.02	012.11	0
012	Starch (%)	2129	40.18	0.1507	39.98	2.293	0.5604	34	0.09	012.04	0
012	Starch (%)	0353	40.21	0.0200	39.98	2.293	0.5604	34	0.10	012.20	0
012	Starch (%)	0693	40.23	0.7700	39.98	2.293	0.5604	34	0.11	012.01	0
012	Starch (%)	0676	40.31	0.0400	39.98	2.293	0.5604	34	0.14	012.01	0
012	Starch (%)	2181	40.33	0.2100	39.98	2.293	0.5604	34	0.15	012.00	0
012	Starch (%)	0265	40.40	0.6000	39.98	2.293	0.5604	34	0.18	012.01	0
012	Starch (%)	0689	40.45	0.7000	39.98	2.293	0.5604	34	0.20	012.00	0
012	Starch (%)	2196	40.58	0.0000	39.98	2.293	0.5604	34	0.26	012.00	0
012	Starch (%)	0683	40.63	0.4900	39.98	2.293	0.5604	34	0.28	012.00	0
012	Starch (%)	0354	40.74	0.0800	39.98	2.293	0.5604	34	0.33	012.00	0
012	Starch (%)	0407	40.74	1.265	39.98	2.293	0.5604	34	0.33	012.03	0
012	Starch (%)	0948	40.92	0.4300	39.98	2.293	0.5604	34	0.41	012.04	0
012	Starch (%)	0164	41.05	0.5000	39.98	2.293	0.5604	34	0.47	012.00	0
012	Starch (%)	0619	41.40	0.2000	39.98	2.293	0.5604	34	0.62	012.00	0
012	Starch (%)	0910	41.40	0.2000	39.98	2.293	0.5604	34	0.62	012.00	0
012	Starch (%)	0878	41.41	0.2700	39.98	2.293	0.5604	34	0.62	012.00	0
012	Starch (%)	2188	41.80	0.1400	39.98	2.293	0.5604	34	0.79	012.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
012	Starch (%)	0889	41.80	0.6200	39.98	2.293	0.5604	34	0.79	012.03	0
012	Starch (%)	2006	42.06	0.6500	39.98	2.293	0.5604	34	0.90	012.00	0
012	Starch (%)	0278	43.10	0.8000	39.98	2.293	0.5604	34	1.36	012.04	0
012	Starch (%)	0297	44.06	0.0700	39.98	2.293	0.5604	34	1.78	012.11	0
012	Starch (%)	0590	46.50	0.0000	39.98	2.293	0.5604	34	2.84	012.11	0
012	Starch (%)	2192	39.48	3.130	39.98	2.293	0.5604	34	-0.22	012.00	1
013	Fat, Acid Pretreat (%)	0618	1.166	0.0078	3.624	0.5298	0.1224	45	-4.64	013.08	0
013	Fat, Acid Pretreat (%)	0083	2.125	0.0100	3.624	0.5298	0.1224	45	-2.83	013.00	0
013	Fat, Acid Pretreat (%)	2073	2.245	0.1100	3.624	0.5298	0.1224	45	-2.60	013.00	0
013	Fat, Acid Pretreat (%)	2146	2.605	0.1500	3.624	0.5298	0.1224	45	-1.92	013.00	0
013	Fat, Acid Pretreat (%)	0689	2.850	0.1000	3.624	0.5298	0.1224	45	-1.46	013.00	0
013	Fat, Acid Pretreat (%)	0683	2.905	0.0900	3.624	0.5298	0.1224	45	-1.36	013.13	0
013	Fat, Acid Pretreat (%)	0939	2.935	0.3500	3.624	0.5298	0.1224	45	-1.30	013.13	0
013	Fat, Acid Pretreat (%)	0036	2.976	0.2207	3.624	0.5298	0.1224	45	-1.22	013.10	0
013	Fat, Acid Pretreat (%)	0297	3.070	0.0000	3.624	0.5298	0.1224	45	-1.05	013.00	0
013	Fat, Acid Pretreat (%)	0968	3.225	0.0300	3.624	0.5298	0.1224	45	-0.75	013.13	0
013	Fat, Acid Pretreat (%)	0353	3.290	0.1200	3.624	0.5298	0.1224	45	-0.63	013.10	0
013	Fat, Acid Pretreat (%)	0652	3.450	0.1000	3.624	0.5298	0.1224	45	-0.33	013.00	0
013	Fat, Acid Pretreat (%)	0910	3.490	0.3600	3.624	0.5298	0.1224	45	-0.25	013.00	0
013	Fat, Acid Pretreat (%)	2076	3.495	0.0011	3.624	0.5298	0.1224	45	-0.24	013.00	0
013	Fat, Acid Pretreat (%)	2259	3.538	0.0960	3.624	0.5298	0.1224	45	-0.16	013.02	0
013	Fat, Acid Pretreat (%)	0853	3.550	0.1000	3.624	0.5298	0.1224	45	-0.14	013.02	0
013	Fat, Acid Pretreat (%)	2196	3.550	0.0000	3.624	0.5298	0.1224	45	-0.14	013.10	0
013	Fat, Acid Pretreat (%)	0541	3.570	0.0600	3.624	0.5298	0.1224	45	-0.10	013.00	0
013	Fat, Acid Pretreat (%)	0948	3.580	0.0600	3.624	0.5298	0.1224	45	-0.08	013.00	0
013	Fat, Acid Pretreat (%)	0309	3.592	0.0158	3.624	0.5298	0.1224	45	-0.06	013.00	0
013	Fat, Acid Pretreat (%)	0229	3.595	0.2700	3.624	0.5298	0.1224	45	-0.05	013.00	0
013	Fat, Acid Pretreat (%)	0148	3.595	0.2300	3.624	0.5298	0.1224	45	-0.05	013.02	0
013	Fat, Acid Pretreat (%)	0904	3.595	0.3300	3.624	0.5298	0.1224	45	-0.05	013.13	0
013	Fat, Acid Pretreat (%)	0876	3.600	0.0000	3.624	0.5298	0.1224	45	-0.05	013.13	0
013	Fat, Acid Pretreat (%)	0098	3.610	0.0600	3.624	0.5298	0.1224	45	-0.03	013.02	0
013	Fat, Acid Pretreat (%)	0914	3.660	0.0600	3.624	0.5298	0.1224	45	0.07	013.02	0
013	Fat, Acid Pretreat (%)	0870	3.688	0.1621	3.624	0.5298	0.1224	45	0.12	013.02	0
013	Fat, Acid Pretreat (%)	0265	3.745	0.2500	3.624	0.5298	0.1224	45	0.23	013.13	0
013	Fat, Acid Pretreat (%)	0843	3.790	0.4200	3.624	0.5298	0.1224	45	0.31	013.13	0
013	Fat, Acid Pretreat (%)	0354	3.840	0.0600	3.624	0.5298	0.1224	45	0.41	013.02	0
013	Fat, Acid Pretreat (%)	0074	3.850	0.2400	3.624	0.5298	0.1224	45	0.43	013.02	0
013	Fat, Acid Pretreat (%)	2304	3.935	0.0300	3.624	0.5298	0.1224	45	0.59	013.00	0
013	Fat, Acid Pretreat (%)	0921	3.940	0.0200	3.624	0.5298	0.1224	45	0.60	013.02	0
013	Fat, Acid Pretreat (%)	0643	4.020	0.0600	3.624	0.5298	0.1224	45	0.75	013.02	0
013	Fat, Acid Pretreat (%)	0171	4.030	0.1000	3.624	0.5298	0.1224	45	0.77	013.02	0
013	Fat, Acid Pretreat (%)	0553	4.035	0.0700	3.624	0.5298	0.1224	45	0.78	013.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
013	Fat, Acid Pretreat (%)	0581	4.040	0.1600	3.624	0.5298	0.1224	45	0.79	013.13	0
013	Fat, Acid Pretreat (%)	0675	4.115	0.0900	3.624	0.5298	0.1224	45	0.93	013.02	0
013	Fat, Acid Pretreat (%)	0682	4.150	0.0000	3.624	0.5298	0.1224	45	0.99	013.02	0
013	Fat, Acid Pretreat (%)	0676	4.195	0.0300	3.624	0.5298	0.1224	45	1.08	013.02	0
013	Fat, Acid Pretreat (%)	2150	4.235	0.1700	3.624	0.5298	0.1224	45	1.15	013.00	0
013	Fat, Acid Pretreat (%)	0202	4.255	0.3700	3.624	0.5298	0.1224	45	1.19	013.00	0
013	Fat, Acid Pretreat (%)	0164	4.340	0.1000	3.624	0.5298	0.1224	45	1.35	013.02	0
013	Fat, Acid Pretreat (%)	0407	4.421	0.1160	3.624	0.5298	0.1224	45	1.50	013.13	0
013	Fat, Acid Pretreat (%)	0407	4.575	0.1290	3.624	0.5298	0.1224	45	1.79	013.02	0
013	Fat, Acid Pretreat (%)	0242	3.530	0.5800	3.624	0.5298	0.1224	45	-0.18	013.00	1
015	Aluminum (ppm)	0037	56.55	3.100	98.77	11.40	5.035	18	-3.70	015.42	0
015	Aluminum (ppm)	0560	77.40	17.60	98.77	11.40	5.035	18	-1.87	015.52	0
015	Aluminum (ppm)	0171	85.00	2.000	98.77	11.40	5.035	18	-1.21	015.41	0
015	Aluminum (ppm)	0510	91.50	1.000	98.77	11.40	5.035	18	-0.64	015.43	0
015	Aluminum (ppm)	0407	93.83	2.184	98.77	11.40	5.035	18	-0.43	015.41	0
015	Aluminum (ppm)	0297	95.50	1.000	98.77	11.40	5.035	18	-0.29	015.43	0
015	Aluminum (ppm)	0353	96.50	3.000	98.77	11.40	5.035	18	-0.20	015.43	0
015	Aluminum (ppm)	0918	96.67	2.005	98.77	11.40	5.035	18	-0.18	015.43	0
015	Aluminum (ppm)	0425	96.80	5.600	98.77	11.40	5.035	18	-0.17	015.43	0
015	Aluminum (ppm)	0164	98.00	0.0000	98.77	11.40	5.035	18	-0.07	015.41	0
015	Aluminum (ppm)	0345	100.7	4.630	98.77	11.40	5.035	18	0.17	015.43	0
015	Aluminum (ppm)	0169	103.5	21.00	98.77	11.40	5.035	18	0.41	015.43	0
015	Aluminum (ppm)	2033	105.3	0.5000	98.77	11.40	5.035	18	0.57	015.43	0
015	Aluminum (ppm)	0553	105.3	13.40	98.77	11.40	5.035	18	0.57	015.53	0
015	Aluminum (ppm)	2129	105.3	1.200	98.77	11.40	5.035	18	0.57	015.42	0
015	Aluminum (ppm)	0520	109.0	0.0000	98.77	11.40	5.035	18	0.90	015.41	0
015	Aluminum (ppm)	0407	116.0	0.4047	98.77	11.40	5.035	18	1.51	015.43	0
015	Aluminum (ppm)	0939	122.0	12.00	98.77	11.40	5.035	18	2.04	015.53	0
017	Boron (ppm)	0560	6.390	1.080	7.934	1.252	0.3108	19	-1.23	017.52	0
017	Boron (ppm)	0553	6.445	0.1500	7.934	1.252	0.3108	19	-1.19	017.53	0
017	Boron (ppm)	0358	6.745	0.1500	7.934	1.252	0.3108	19	-0.95	017.41	0
017	Boron (ppm)	0693	6.925	0.2600	7.934	1.252	0.3108	19	-0.81	017.42	0
017	Boron (ppm)	0083	7.100	0.0000	7.934	1.252	0.3108	19	-0.67	017.43	0
017	Boron (ppm)	2033	7.150	0.1000	7.934	1.252	0.3108	19	-0.63	017.43	0
017	Boron (ppm)	0229	7.295	0.1700	7.934	1.252	0.3108	19	-0.51	017.41	0
017	Boron (ppm)	0345	7.360	0.0800	7.934	1.252	0.3108	19	-0.46	017.43	0
017	Boron (ppm)	0037	7.500	0.4000	7.934	1.252	0.3108	19	-0.35	017.42	0
017	Boron (ppm)	0045	7.565	0.9700	7.934	1.252	0.3108	19	-0.29	017.42	0
017	Boron (ppm)	0407	7.905	0.2697	7.934	1.252	0.3108	19	-0.02	017.41	0
017	Boron (ppm)	0407	7.960	0.3629	7.934	1.252	0.3108	19	0.02	017.43	0
017	Boron (ppm)	0294	8.725	0.0500	7.934	1.252	0.3108	19	0.63	017.42	0
017	Boron (ppm)	0265	8.800	1.000	7.934	1.252	0.3108	19	0.69	017.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
017	Boron (ppm)	0226	8.910	0.0800	7.934	1.252	0.3108	19	0.78	017.41	0
017	Boron (ppm)	0353	9.000	0.0000	7.934	1.252	0.3108	19	0.85	017.43	0
017	Boron (ppm)	2129	9.352	0.5520	7.934	1.252	0.3108	19	1.13	017.42	0
017	Boron (ppm)	0918	9.985	0.2300	7.934	1.252	0.3108	19	1.64	017.43	0
017	Boron (ppm)	0510	10.00	0.0000	7.934	1.252	0.3108	19	1.65	017.43	0
017	Boron (ppm)	0297	0.0000	0.0000	7.934	1.252	0.3108	19		017.43	4
019	Calcium (%)	0038	1.665	0.0100	2.192	0.1184	0.0490	118	-4.45	019.43	0
019	Calcium (%)	2181	1.825	0.0500	2.192	0.1184	0.0490	118	-3.10	019.00	0
019	Calcium (%)	2129	1.897	0.0240	2.192	0.1184	0.0490	118	-2.49	019.42	0
019	Calcium (%)	2033	1.900	0.0000	2.192	0.1184	0.0490	118	-2.46	019.43	0
019	Calcium (%)	2192	1.915	0.0100	2.192	0.1184	0.0490	118	-2.34	019.43	0
019	Calcium (%)	0175	1.925	0.0300	2.192	0.1184	0.0490	118	-2.25	019.31	0
019	Calcium (%)	2234	1.975	0.0500	2.192	0.1184	0.0490	118	-1.83	019.43	0
019	Calcium (%)	2246	1.980	0.0200	2.192	0.1184	0.0490	118	-1.79	019.31	0
019	Calcium (%)	2246	2.010	0.0000	2.192	0.1184	0.0490	118	-1.54	019.00	0
019	Calcium (%)	0968	2.019	0.0000	2.192	0.1184	0.0490	118	-1.46	019.43	0
019	Calcium (%)	0263	2.027	0.0500	2.192	0.1184	0.0490	118	-1.39	019.42	0
019	Calcium (%)	0278	2.040	0.0200	2.192	0.1184	0.0490	118	-1.28	019.42	0
019	Calcium (%)	0089	2.040	0.0000	2.192	0.1184	0.0490	118	-1.28	019.43	0
019	Calcium (%)	0357	2.044	0.0987	2.192	0.1184	0.0490	118	-1.25	019.42	0
019	Calcium (%)	0242	2.045	0.1500	2.192	0.1184	0.0490	118	-1.24	019.99	0
019	Calcium (%)	2188	2.055	0.0500	2.192	0.1184	0.0490	118	-1.15	019.08	0
019	Calcium (%)	0144	2.055	0.0900	2.192	0.1184	0.0490	118	-1.15	019.41	0
019	Calcium (%)	0036	2.060	0.0302	2.192	0.1184	0.0490	118	-1.11	019.43	0
019	Calcium (%)	0529	2.070	0.0400	2.192	0.1184	0.0490	118	-1.03	019.31	0
019	Calcium (%)	0154	2.087	0.0190	2.192	0.1184	0.0490	118	-0.89	019.52	0
019	Calcium (%)	0265	2.090	0.0400	2.192	0.1184	0.0490	118	-0.86	019.42	0
019	Calcium (%)	2146	2.095	0.1100	2.192	0.1184	0.0490	118	-0.82	019.31	0
019	Calcium (%)	0169	2.095	0.0700	2.192	0.1184	0.0490	118	-0.82	019.32	0
019	Calcium (%)	0675	2.095	0.0100	2.192	0.1184	0.0490	118	-0.82	019.43	0
019	Calcium (%)	0918	2.097	0.0040	2.192	0.1184	0.0490	118	-0.80	019.43	0
019	Calcium (%)	2295	2.105	0.0100	2.192	0.1184	0.0490	118	-0.73	019.00	0
019	Calcium (%)	0674	2.110	0.0000	2.192	0.1184	0.0490	118	-0.69	019.31	0
019	Calcium (%)	2006	2.110	0.1740	2.192	0.1184	0.0490	118	-0.69	019.00	0
019	Calcium (%)	0910	2.110	0.0600	2.192	0.1184	0.0490	118	-0.69	019.41	0
019	Calcium (%)	0693	2.116	0.0440	2.192	0.1184	0.0490	118	-0.64	019.42	0
019	Calcium (%)	0123	2.120	0.0000	2.192	0.1184	0.0490	118	-0.61	019.41	0
019	Calcium (%)	2006	2.123	0.0940	2.192	0.1184	0.0490	118	-0.58	019.09	0
019	Calcium (%)	0142	2.123	0.0092	2.192	0.1184	0.0490	118	-0.58	019.31	0
019	Calcium (%)	0590	2.125	0.0700	2.192	0.1184	0.0490	118	-0.56	019.08	0
019	Calcium (%)	2012	2.135	0.0100	2.192	0.1184	0.0490	118	-0.48	019.41	0
019	Calcium (%)	2259	2.138	0.0220	2.192	0.1184	0.0490	118	-0.45	019.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
019	Calcium (%)	0108	2.140	0.1400	2.192	0.1184	0.0490	118	-0.44	019.31	0
019	Calcium (%)	0171	2.140	0.0200	2.192	0.1184	0.0490	118	-0.44	019.41	0
019	Calcium (%)	0148	2.145	0.0100	2.192	0.1184	0.0490	118	-0.39	019.41	0
019	Calcium (%)	0354	2.145	0.0300	2.192	0.1184	0.0490	118	-0.39	019.41	0
019	Calcium (%)	0572	2.145	0.1700	2.192	0.1184	0.0490	118	-0.39	019.53	0
019	Calcium (%)	0590	2.155	0.1300	2.192	0.1184	0.0490	118	-0.31	019.99	0
019	Calcium (%)	2009	2.158	0.0045	2.192	0.1184	0.0490	118	-0.29	019.08	0
019	Calcium (%)	0656	2.160	0.0000	2.192	0.1184	0.0490	118	-0.27	019.31	0
019	Calcium (%)	0407	2.162	0.0943	2.192	0.1184	0.0490	118	-0.25	019.41	0
019	Calcium (%)	0035	2.165	0.0700	2.192	0.1184	0.0490	118	-0.23	019.42	0
019	Calcium (%)	0692	2.165	0.0100	2.192	0.1184	0.0490	118	-0.23	019.42	0
019	Calcium (%)	0298	2.170	0.0200	2.192	0.1184	0.0490	118	-0.18	019.41	0
019	Calcium (%)	0083	2.170	0.0400	2.192	0.1184	0.0490	118	-0.18	019.43	0
019	Calcium (%)	0297	2.170	0.1600	2.192	0.1184	0.0490	118	-0.18	019.43	0
019	Calcium (%)	0505	2.170	0.0200	2.192	0.1184	0.0490	118	-0.18	019.43	0
019	Calcium (%)	0037	2.176	0.0100	2.192	0.1184	0.0490	118	-0.13	019.42	0
019	Calcium (%)	0914	2.178	0.1764	2.192	0.1184	0.0490	118	-0.12	019.43	0
019	Calcium (%)	0939	2.180	0.0400	2.192	0.1184	0.0490	118	-0.10	019.31	0
019	Calcium (%)	0881	2.181	0.0550	2.192	0.1184	0.0490	118	-0.09	019.00	0
019	Calcium (%)	0345	2.185	0.0300	2.192	0.1184	0.0490	118	-0.06	019.43	0
019	Calcium (%)	0893	2.190	0.0000	2.192	0.1184	0.0490	118	-0.01	019.00	0
019	Calcium (%)	0676	2.190	0.0800	2.192	0.1184	0.0490	118	-0.01	019.99	0
019	Calcium (%)	0878	2.190	0.0200	2.192	0.1184	0.0490	118	-0.01	019.00	0
019	Calcium (%)	2144	2.190	0.0200	2.192	0.1184	0.0490	118	-0.01	019.00	0
019	Calcium (%)	0066	2.195	0.0350	2.192	0.1184	0.0490	118	0.02	019.41	0
019	Calcium (%)	0897	2.195	0.0100	2.192	0.1184	0.0490	118	0.03	019.00	0
019	Calcium (%)	0425	2.195	0.0700	2.192	0.1184	0.0490	118	0.03	019.43	0
019	Calcium (%)	0510	2.200	0.0000	2.192	0.1184	0.0490	118	0.07	019.43	0
019	Calcium (%)	0553	2.200	0.0200	2.192	0.1184	0.0490	118	0.07	019.53	0
019	Calcium (%)	0036	2.205	0.0670	2.192	0.1184	0.0490	118	0.11	019.03	0
019	Calcium (%)	2062	2.208	0.0252	2.192	0.1184	0.0490	118	0.14	019.00	0
019	Calcium (%)	2196	2.210	0.0000	2.192	0.1184	0.0490	118	0.15	019.31	0
019	Calcium (%)	0019	2.210	0.0200	2.192	0.1184	0.0490	118	0.15	019.41	0
019	Calcium (%)	0848	2.210	0.0400	2.192	0.1184	0.0490	118	0.15	019.41	0
019	Calcium (%)	0878	2.210	0.0200	2.192	0.1184	0.0490	118	0.15	019.41	0
019	Calcium (%)	0047	2.220	0.1000	2.192	0.1184	0.0490	118	0.24	019.52	0
019	Calcium (%)	2196	2.220	0.0000	2.192	0.1184	0.0490	118	0.24	019.08	0
019	Calcium (%)	0870	2.221	0.0670	2.192	0.1184	0.0490	118	0.24	019.42	0
019	Calcium (%)	2022	2.225	0.0300	2.192	0.1184	0.0490	118	0.28	019.00	0
019	Calcium (%)	0536	2.240	0.0400	2.192	0.1184	0.0490	118	0.41	019.31	0
019	Calcium (%)	0035	2.240	0.0000	2.192	0.1184	0.0490	118	0.41	019.43	0
019	Calcium (%)	2152	2.240	0.0200	2.192	0.1184	0.0490	118	0.41	019.44	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
019	Calcium (%)	0885	2.246	0.0097	2.192	0.1184	0.0490	118	0.46	019.08	0
019	Calcium (%)	2302	2.250	0.0000	2.192	0.1184	0.0490	118	0.49	019.00	0
019	Calcium (%)	0948	2.253	0.0420	2.192	0.1184	0.0490	118	0.52	019.31	0
019	Calcium (%)	0683	2.255	0.0700	2.192	0.1184	0.0490	118	0.53	019.00	0
019	Calcium (%)	0042	2.255	0.0100	2.192	0.1184	0.0490	118	0.53	019.42	0
019	Calcium (%)	0560	2.255	0.2100	2.192	0.1184	0.0490	118	0.53	019.52	0
019	Calcium (%)	0689	2.260	0.0200	2.192	0.1184	0.0490	118	0.58	019.08	0
019	Calcium (%)	0045	2.260	0.2000	2.192	0.1184	0.0490	118	0.58	019.42	0
019	Calcium (%)	0682	2.260	0.0000	2.192	0.1184	0.0490	118	0.58	019.43	0
019	Calcium (%)	0939	2.260	0.0200	2.192	0.1184	0.0490	118	0.58	019.53	0
019	Calcium (%)	0563	2.264	0.0218	2.192	0.1184	0.0490	118	0.61	019.31	0
019	Calcium (%)	2109	2.265	0.1340	2.192	0.1184	0.0490	118	0.62	019.00	0
019	Calcium (%)	0407	2.266	0.0717	2.192	0.1184	0.0490	118	0.63	019.43	0
019	Calcium (%)	0294	2.270	0.0400	2.192	0.1184	0.0490	118	0.66	019.42	0
019	Calcium (%)	0164	2.270	0.0400	2.192	0.1184	0.0490	118	0.66	019.44	0
019	Calcium (%)	0001	2.276	0.0280	2.192	0.1184	0.0490	118	0.71	019.31	0
019	Calcium (%)	0010	2.280	0.0400	2.192	0.1184	0.0490	118	0.75	019.43	0
019	Calcium (%)	0004	2.290	0.0000	2.192	0.1184	0.0490	118	0.83	019.41	0
019	Calcium (%)	0229	2.295	0.0300	2.192	0.1184	0.0490	118	0.87	019.41	0
019	Calcium (%)	0868	2.295	0.0700	2.192	0.1184	0.0490	118	0.87	019.99	0
019	Calcium (%)	0870	2.297	0.0390	2.192	0.1184	0.0490	118	0.89	019.43	0
019	Calcium (%)	0687	2.300	0.2000	2.192	0.1184	0.0490	118	0.91	019.31	0
019	Calcium (%)	0358	2.300	0.0800	2.192	0.1184	0.0490	118	0.91	019.41	0
019	Calcium (%)	0868	2.305	0.0100	2.192	0.1184	0.0490	118	0.96	019.31	0
019	Calcium (%)	0003	2.320	0.1600	2.192	0.1184	0.0490	118	1.08	019.41	0
019	Calcium (%)	0100	2.320	0.1600	2.192	0.1184	0.0490	118	1.08	019.99	0
019	Calcium (%)	0904	2.340	0.1600	2.192	0.1184	0.0490	118	1.25	019.31	0
019	Calcium (%)	0190	2.365	0.0100	2.192	0.1184	0.0490	118	1.46	019.42	0
019	Calcium (%)	0619	2.375	0.0100	2.192	0.1184	0.0490	118	1.55	019.41	0
019	Calcium (%)	0226	2.380	0.0200	2.192	0.1184	0.0490	118	1.59	019.41	0
019	Calcium (%)	0202	2.385	0.0100	2.192	0.1184	0.0490	118	1.63	019.43	0
019	Calcium (%)	0598	2.392	0.0790	2.192	0.1184	0.0490	118	1.69	019.41	0
019	Calcium (%)	0009	2.425	0.0500	2.192	0.1184	0.0490	118	1.97	019.42	0
019	Calcium (%)	2302	2.440	0.0000	2.192	0.1184	0.0490	118	2.10	019.99	0
019	Calcium (%)	0098	2.445	0.0500	2.192	0.1184	0.0490	118	2.14	019.41	0
019	Calcium (%)	0353	2.450	0.0200	2.192	0.1184	0.0490	118	2.18	019.43	0
019	Calcium (%)	0098	2.485	0.1300	2.192	0.1184	0.0490	118	2.48	019.43	0
019	Calcium (%)	0520	2.495	0.0100	2.192	0.1184	0.0490	118	2.56	019.41	0
019	Calcium (%)	0964	2.502	0.0786	2.192	0.1184	0.0490	118	2.62	019.43	0
019	Calcium (%)	0726	2.580	0.0600	2.192	0.1184	0.0490	118	3.28	019.42	0
019	Calcium (%)	0038	1.885	0.2900	2.192	0.1184	0.0490	118	-2.59	019.31	1
019	Calcium (%)	0074	2.335	0.3100	2.192	0.1184	0.0490	118	1.21	019.41	1

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
021	Cobalt (ppm)	0689	0.5500	0.1000	0.9901	0.3076	0.0725	21	-1.43	021.31	0
021	Cobalt (ppm)	0693	0.5750	0.2840	0.9901	0.3076	0.0725	21	-1.35	021.42	0
021	Cobalt (ppm)	0619	0.6810	0.0120	0.9901	0.3076	0.0725	21	-1.00	021.41	0
021	Cobalt (ppm)	0910	0.6950	0.0500	0.9901	0.3076	0.0725	21	-0.96	021.52	0
021	Cobalt (ppm)	0918	0.7306	0.0162	0.9901	0.3076	0.0725	21	-0.84	021.53	0
021	Cobalt (ppm)	0345	0.7650	0.0700	0.9901	0.3076	0.0725	21	-0.73	021.43	0
021	Cobalt (ppm)	0171	0.7690	0.0100	0.9901	0.3076	0.0725	21	-0.72	021.41	0
021	Cobalt (ppm)	0038	0.8360	0.0020	0.9901	0.3076	0.0725	21	-0.50	021.43	0
021	Cobalt (ppm)	0169	0.8550	0.1100	0.9901	0.3076	0.0725	21	-0.44	021.43	0
021	Cobalt (ppm)	0560	0.8650	0.0100	0.9901	0.3076	0.0725	21	-0.41	021.52	0
021	Cobalt (ppm)	0553	1.080	0.0400	0.9901	0.3076	0.0725	21	0.29	021.53	0
021	Cobalt (ppm)	0964	1.084	0.0211	0.9901	0.3076	0.0725	21	0.30	021.43	0
021	Cobalt (ppm)	0510	1.165	0.0100	0.9901	0.3076	0.0725	21	0.57	021.43	0
021	Cobalt (ppm)	0045	1.195	0.1900	0.9901	0.3076	0.0725	21	0.67	021.42	0
021	Cobalt (ppm)	0148	1.215	0.1100	0.9901	0.3076	0.0725	21	0.73	021.41	0
021	Cobalt (ppm)	0939	1.225	0.1300	0.9901	0.3076	0.0725	21	0.76	021.53	0
021	Cobalt (ppm)	0164	1.250	0.1000	0.9901	0.3076	0.0725	21	0.84	021.31	0
021	Cobalt (ppm)	0939	1.255	0.0900	0.9901	0.3076	0.0725	21	0.86	021.31	0
021	Cobalt (ppm)	0407	1.283	0.0094	0.9901	0.3076	0.0725	21	0.95	021.41	0
021	Cobalt (ppm)	2033	1.300	0.0400	0.9901	0.3076	0.0725	21	1.01	021.43	0
021	Cobalt (ppm)	0407	1.419	0.1180	0.9901	0.3076	0.0725	21	1.39	021.43	0
021	Cobalt (ppm)	0353	1.262	1.297	0.9901	0.3076	0.0725	21	0.88	021.43	1
021	Cobalt (ppm)	0682	< 1		0.9901	0.3076	0.0725	21		021.43	5
022	Copper (ppm)	2234	31.32	3.780	44.29	3.112	1.792	83	-4.17	022.43	0
022	Copper (ppm)	2302	33.00	0.0000	44.29	3.112	1.792	83	-3.63	022.99	0
022	Copper (ppm)	0536	34.85	0.0000	44.29	3.112	1.792	83	-3.03	022.31	0
022	Copper (ppm)	0619	34.85	0.3000	44.29	3.112	1.792	83	-3.03	022.41	0
022	Copper (ppm)	0171	36.33	0.7400	44.29	3.112	1.792	83	-2.56	022.41	0
022	Copper (ppm)	0656	37.20	0.0000	44.29	3.112	1.792	83	-2.28	022.31	0
022	Copper (ppm)	0590	37.50	1.800	44.29	3.112	1.792	83	-2.18	022.99	0
022	Copper (ppm)	0047	37.77	3.430	44.29	3.112	1.792	83	-2.10	022.52	0
022	Copper (ppm)	0003	38.50	1.000	44.29	3.112	1.792	83	-1.86	022.41	0
022	Copper (ppm)	0148	38.50	0.2000	44.29	3.112	1.792	83	-1.86	022.41	0
022	Copper (ppm)	0520	40.00	2.000	44.29	3.112	1.792	83	-1.38	022.41	0
022	Copper (ppm)	0004	40.44	1.640	44.29	3.112	1.792	83	-1.24	022.41	0
022	Copper (ppm)	0910	40.50	5.000	44.29	3.112	1.792	83	-1.22	022.41	0
022	Copper (ppm)	0553	41.50	3.200	44.29	3.112	1.792	83	-0.90	022.53	0
022	Copper (ppm)	0563	41.98	1.353	44.29	3.112	1.792	83	-0.74	022.31	0
022	Copper (ppm)	0870	42.01	2.910	44.29	3.112	1.792	83	-0.74	022.42	0
022	Copper (ppm)	2188	42.07	0.8000	44.29	3.112	1.792	83	-0.71	022.31	0
022	Copper (ppm)	0939	42.52	0.1000	44.29	3.112	1.792	83	-0.57	022.31	0
022	Copper (ppm)	0226	42.53	1.200	44.29	3.112	1.792	83	-0.57	022.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
022	Copper (ppm)	0045	42.55	2.500	44.29	3.112	1.792	83	-0.56	022.42	0
022	Copper (ppm)	0529	42.70	4.600	44.29	3.112	1.792	83	-0.51	022.31	0
022	Copper (ppm)	0357	42.72	0.6890	44.29	3.112	1.792	83	-0.51	022.42	0
022	Copper (ppm)	0038	42.75	0.5000	44.29	3.112	1.792	83	-0.50	022.32	0
022	Copper (ppm)	0229	42.79	0.5000	44.29	3.112	1.792	83	-0.48	022.41	0
022	Copper (ppm)	0353	43.00	4.000	44.29	3.112	1.792	83	-0.42	022.43	0
022	Copper (ppm)	0100	43.00	2.000	44.29	3.112	1.792	83	-0.42	022.99	0
022	Copper (ppm)	0407	43.07	4.050	44.29	3.112	1.792	83	-0.39	022.41	0
022	Copper (ppm)	0354	43.12	0.2300	44.29	3.112	1.792	83	-0.38	022.41	0
022	Copper (ppm)	0918	43.18	2.690	44.29	3.112	1.792	83	-0.36	022.43	0
022	Copper (ppm)	2012	43.21	2.540	44.29	3.112	1.792	83	-0.35	022.41	0
022	Copper (ppm)	0358	43.43	3.580	44.29	3.112	1.792	83	-0.28	022.41	0
022	Copper (ppm)	0848	43.45	0.4800	44.29	3.112	1.792	83	-0.27	022.41	0
022	Copper (ppm)	0572	43.50	0.6000	44.29	3.112	1.792	83	-0.25	022.53	0
022	Copper (ppm)	0242	43.50	5.000	44.29	3.112	1.792	83	-0.25	022.99	0
022	Copper (ppm)	0948	43.66	0.1710	44.29	3.112	1.792	83	-0.20	022.33	0
022	Copper (ppm)	0202	43.79	0.6200	44.29	3.112	1.792	83	-0.16	022.43	0
022	Copper (ppm)	0164	44.00	2.000	44.29	3.112	1.792	83	-0.09	022.44	0
022	Copper (ppm)	2129	44.05	0.0900	44.29	3.112	1.792	83	-0.08	022.42	0
022	Copper (ppm)	0169	44.10	2.000	44.29	3.112	1.792	83	-0.06	022.43	0
022	Copper (ppm)	0186	44.35	2.700	44.29	3.112	1.792	83	0.02	022.52	0
022	Copper (ppm)	0675	44.48	1.850	44.29	3.112	1.792	83	0.06	022.43	0
022	Copper (ppm)	0074	44.50	1.000	44.29	3.112	1.792	83	0.07	022.41	0
022	Copper (ppm)	0878	44.50	1.000	44.29	3.112	1.792	83	0.07	022.41	0
022	Copper (ppm)	0968	44.50	1.000	44.29	3.112	1.792	83	0.07	022.43	0
022	Copper (ppm)	0560	44.60	0.4000	44.29	3.112	1.792	83	0.10	022.52	0
022	Copper (ppm)	0263	44.62	0.7090	44.29	3.112	1.792	83	0.10	022.42	0
022	Copper (ppm)	0278	44.75	2.300	44.29	3.112	1.792	83	0.15	022.42	0
022	Copper (ppm)	0175	45.00	2.000	44.29	3.112	1.792	83	0.23	022.31	0
022	Copper (ppm)	0083	45.00	2.000	44.29	3.112	1.792	83	0.23	022.43	0
022	Copper (ppm)	0693	45.01	5.041	44.29	3.112	1.792	83	0.23	022.42	0
022	Copper (ppm)	0868	45.05	0.9000	44.29	3.112	1.792	83	0.24	022.31	0
022	Copper (ppm)	2196	45.20	0.0000	44.29	3.112	1.792	83	0.29	022.31	0
022	Copper (ppm)	0037	45.30	1.200	44.29	3.112	1.792	83	0.32	022.42	0
022	Copper (ppm)	0692	45.30	2.200	44.29	3.112	1.792	83	0.32	022.42	0
022	Copper (ppm)	0407	45.35	1.854	44.29	3.112	1.792	83	0.34	022.43	0
022	Copper (ppm)	0726	45.38	0.1500	44.29	3.112	1.792	83	0.35	022.42	0
022	Copper (ppm)	0505	45.44	3.940	44.29	3.112	1.792	83	0.37	022.43	0
022	Copper (ppm)	0098	45.47	2.650	44.29	3.112	1.792	83	0.38	022.41	0
022	Copper (ppm)	0425	45.60	3.600	44.29	3.112	1.792	83	0.42	022.43	0
022	Copper (ppm)	0939	45.75	0.4900	44.29	3.112	1.792	83	0.47	022.53	0
022	Copper (ppm)	0345	45.83	0.7500	44.29	3.112	1.792	83	0.49	022.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
022	Copper (ppm)	0682	45.90	0.0000	44.29	3.112	1.792	83	0.52	022.43	0
022	Copper (ppm)	0019	46.37	1.750	44.29	3.112	1.792	83	0.67	022.41	0
022	Copper (ppm)	2192	46.48	0.9900	44.29	3.112	1.792	83	0.70	022.43	0
022	Copper (ppm)	0510	46.50	1.000	44.29	3.112	1.792	83	0.71	022.43	0
022	Copper (ppm)	0098	46.66	2.260	44.29	3.112	1.792	83	0.76	022.43	0
022	Copper (ppm)	0870	46.99	1.450	44.29	3.112	1.792	83	0.86	022.43	0
022	Copper (ppm)	0009	47.00	2.000	44.29	3.112	1.792	83	0.87	022.42	0
022	Copper (ppm)	0010	47.05	3.500	44.29	3.112	1.792	83	0.89	022.33	0
022	Copper (ppm)	0265	47.50	1.000	44.29	3.112	1.792	83	1.03	022.42	0
022	Copper (ppm)	0914	47.60	5.200	44.29	3.112	1.792	83	1.06	022.43	0
022	Copper (ppm)	0186	47.65	3.300	44.29	3.112	1.792	83	1.08	022.42	0
022	Copper (ppm)	0066	47.68	0.6400	44.29	3.112	1.792	83	1.09	022.44	0
022	Copper (ppm)	0042	47.85	0.5000	44.29	3.112	1.792	83	1.14	022.42	0
022	Copper (ppm)	0598	48.08	1.850	44.29	3.112	1.792	83	1.22	022.41	0
022	Copper (ppm)	2033	48.45	1.300	44.29	3.112	1.792	83	1.34	022.43	0
022	Copper (ppm)	0035	48.50	1.000	44.29	3.112	1.792	83	1.35	022.43	0
022	Copper (ppm)	0190	49.25	0.7700	44.29	3.112	1.792	83	1.59	022.42	0
022	Copper (ppm)	0035	49.50	1.000	44.29	3.112	1.792	83	1.67	022.42	0
022	Copper (ppm)	0297	49.50	5.000	44.29	3.112	1.792	83	1.67	022.43	0
022	Copper (ppm)	0038	50.10	2.000	44.29	3.112	1.792	83	1.87	022.43	0
022	Copper (ppm)	0964	51.03	4.120	44.29	3.112	1.792	83	2.17	022.43	0
022	Copper (ppm)	0294	53.64	2.080	44.29	3.112	1.792	83	3.00	022.42	0
023	Fluorine (ppm)	2033	23.00	1.000				1		023.01	0
024	Iodine (ppm)	0171	3.450	0.1000				1		024.03	0
025	Iron (ppm)	0536	14.43	0.6100	641.2	48.63	14.06	77	-12.89	025.31	0
025	Iron (ppm)	0278	203.5	27.00	641.2	48.63	14.06	77	-9.00	025.42	0
025	Iron (ppm)	0726	322.4	3.900	641.2	48.63	14.06	77	-6.56	025.42	0
025	Iron (ppm)	0265	340.5	19.00	641.2	48.63	14.06	77	-6.18	025.42	0
025	Iron (ppm)	0904	391.0	4.750	641.2	48.63	14.06	77	-5.14	025.33	0
025	Iron (ppm)	0047	479.3	0.2000	641.2	48.63	14.06	77	-3.33	025.52	0
025	Iron (ppm)	0345	479.5	13.00	641.2	48.63	14.06	77	-3.32	025.43	0
025	Iron (ppm)	2234	505.4	31.21	641.2	48.63	14.06	77	-2.79	025.43	0
025	Iron (ppm)	2178	506.5	51.00	641.2	48.63	14.06	77	-2.77	025.33	0
025	Iron (ppm)	0619	540.9	3.694	641.2	48.63	14.06	77	-2.06	025.41	0
025	Iron (ppm)	0226	542.3	21.24	641.2	48.63	14.06	77	-2.03	025.41	0
025	Iron (ppm)	0190	543.7	6.500	641.2	48.63	14.06	77	-2.01	025.42	0
025	Iron (ppm)	0294	560.6	0.9400	641.2	48.63	14.06	77	-1.66	025.42	0
025	Iron (ppm)	0689	573.8	3.600	641.2	48.63	14.06	77	-1.39	025.31	0
025	Iron (ppm)	0042	603.5	21.00	641.2	48.63	14.06	77	-0.78	025.42	0
025	Iron (ppm)	0656	605.6	0.0000	641.2	48.63	14.06	77	-0.73	025.31	0
025	Iron (ppm)	0263	610.9	8.448	641.2	48.63	14.06	77	-0.62	025.42	0
025	Iron (ppm)	0123	613.5	27.00	641.2	48.63	14.06	77	-0.57	025.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
025	Iron (ppm)	0675	614.2	0.0100	641.2	48.63	14.06	77	-0.56	025.43	0
025	Iron (ppm)	0354	616.7	2.080	641.2	48.63	14.06	77	-0.50	025.41	0
025	Iron (ppm)	0353	617.0	10.00	641.2	48.63	14.06	77	-0.50	025.43	0
025	Iron (ppm)	0407	620.0	2.441	641.2	48.63	14.06	77	-0.44	025.41	0
025	Iron (ppm)	2129	620.8	8.300	641.2	48.63	14.06	77	-0.42	025.42	0
025	Iron (ppm)	0003	622.5	43.00	641.2	48.63	14.06	77	-0.38	025.41	0
025	Iron (ppm)	0169	624.5	17.00	641.2	48.63	14.06	77	-0.34	025.43	0
025	Iron (ppm)	0553	625.0	12.00	641.2	48.63	14.06	77	-0.33	025.53	0
025	Iron (ppm)	2302	633.0	0.0000	641.2	48.63	14.06	77	-0.17	025.99	0
025	Iron (ppm)	2146	633.2	3.200	641.2	48.63	14.06	77	-0.16	025.31	0
025	Iron (ppm)	0692	634.0	36.00	641.2	48.63	14.06	77	-0.15	025.42	0
025	Iron (ppm)	0968	634.0	6.000	641.2	48.63	14.06	77	-0.15	025.43	0
025	Iron (ppm)	0175	639.0	6.000	641.2	48.63	14.06	77	-0.05	025.31	0
025	Iron (ppm)	0242	639.0	46.00	641.2	48.63	14.06	77	-0.05	025.99	0
025	Iron (ppm)	2196	639.5	0.0000	641.2	48.63	14.06	77	-0.03	025.31	0
025	Iron (ppm)	2033	640.0	0.0000	641.2	48.63	14.06	77	-0.02	025.43	0
025	Iron (ppm)	0529	642.0	13.70	641.2	48.63	14.06	77	0.02	025.31	0
025	Iron (ppm)	0001	642.5	17.40	641.2	48.63	14.06	77	0.03	025.31	0
025	Iron (ppm)	0037	645.4	24.30	641.2	48.63	14.06	77	0.09	025.42	0
025	Iron (ppm)	0693	647.2	1.714	641.2	48.63	14.06	77	0.12	025.42	0
025	Iron (ppm)	0910	650.0	20.00	641.2	48.63	14.06	77	0.18	025.41	0
025	Iron (ppm)	2192	652.9	45.80	641.2	48.63	14.06	77	0.24	025.43	0
025	Iron (ppm)	0563	653.5	13.00	641.2	48.63	14.06	77	0.25	025.31	0
025	Iron (ppm)	0425	653.7	32.20	641.2	48.63	14.06	77	0.26	025.43	0
025	Iron (ppm)	0870	656.0	32.00	641.2	48.63	14.06	77	0.30	025.42	0
025	Iron (ppm)	0098	658.3	22.50	641.2	48.63	14.06	77	0.35	025.41	0
025	Iron (ppm)	0083	659.0	8.000	641.2	48.63	14.06	77	0.37	025.43	0
025	Iron (ppm)	2188	659.9	1.240	641.2	48.63	14.06	77	0.38	025.31	0
025	Iron (ppm)	0066	662.3	9.440	641.2	48.63	14.06	77	0.43	025.41	0
025	Iron (ppm)	0918	662.6	6.460	641.2	48.63	14.06	77	0.44	025.43	0
025	Iron (ppm)	2144	663.9	2.770	641.2	48.63	14.06	77	0.47	025.99	0
025	Iron (ppm)	0297	664.0	42.00	641.2	48.63	14.06	77	0.47	025.43	0
025	Iron (ppm)	2012	666.2	10.76	641.2	48.63	14.06	77	0.51	025.41	0
025	Iron (ppm)	0520	669.0	10.00	641.2	48.63	14.06	77	0.57	025.41	0
025	Iron (ppm)	0878	669.0	14.00	641.2	48.63	14.06	77	0.57	025.41	0
025	Iron (ppm)	0019	669.1	33.90	641.2	48.63	14.06	77	0.57	025.41	0
025	Iron (ppm)	0505	670.2	9.000	641.2	48.63	14.06	77	0.60	025.31	0
025	Iron (ppm)	0682	670.5	0.0000	641.2	48.63	14.06	77	0.60	025.43	0
025	Iron (ppm)	0939	670.7	2.020	641.2	48.63	14.06	77	0.61	025.31	0
025	Iron (ppm)	0004	671.7	1.650	641.2	48.63	14.06	77	0.63	025.41	0
025	Iron (ppm)	0510	672.0	4.000	641.2	48.63	14.06	77	0.63	025.43	0
025	Iron (ppm)	0074	672.5	1.000	641.2	48.63	14.06	77	0.64	025.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
025	Iron (ppm)	0171	673.1	0.3000	641.2	48.63	14.06	77	0.66	025.41	0
025	Iron (ppm)	0045	675.0	12.00	641.2	48.63	14.06	77	0.70	025.42	0
025	Iron (ppm)	0914	675.8	37.30	641.2	48.63	14.06	77	0.71	025.43	0
025	Iron (ppm)	0148	679.0	6.000	641.2	48.63	14.06	77	0.78	025.41	0
025	Iron (ppm)	0098	679.8	9.400	641.2	48.63	14.06	77	0.79	025.43	0
025	Iron (ppm)	0100	684.5	9.000	641.2	48.63	14.06	77	0.89	025.99	0
025	Iron (ppm)	0939	686.0	4.000	641.2	48.63	14.06	77	0.92	025.53	0
025	Iron (ppm)	0598	687.4	11.10	641.2	48.63	14.06	77	0.95	025.41	0
025	Iron (ppm)	0407	687.5	5.203	641.2	48.63	14.06	77	0.95	025.43	0
025	Iron (ppm)	0358	689.1	42.58	641.2	48.63	14.06	77	0.98	025.41	0
025	Iron (ppm)	0870	690.4	23.90	641.2	48.63	14.06	77	1.01	025.43	0
025	Iron (ppm)	0164	696.5	1.000	641.2	48.63	14.06	77	1.14	025.41	0
025	Iron (ppm)	0868	698.5	25.00	641.2	48.63	14.06	77	1.18	025.31	0
025	Iron (ppm)	0229	706.4	19.80	641.2	48.63	14.06	77	1.34	025.41	0
025	Iron (ppm)	0035	720.0	12.00	641.2	48.63	14.06	77	1.62	025.42	0
025	Iron (ppm)	0035	733.0	0.0000	641.2	48.63	14.06	77	1.89	025.43	0
025	Iron (ppm)	0038	736.0	50.00	641.2	48.63	14.06	77	1.95	025.43	0
025	Iron (ppm)	0560	429.1	171.1	641.2	48.63	14.06	77	-4.36	025.52	1
025	Iron (ppm)	0964	725.2	102.0	641.2	48.63	14.06	77	1.73	025.43	1
025	Iron (ppm)	0038	1,415	114.0	641.2	48.63	14.06	77	15.91	025.31	1
027	Magnesium (%)	0019	0.1850	0.0100	0.2066	0.0103	0.0055	72	-2.10	027.41	0
027	Magnesium (%)	0353	0.1850	0.0100	0.2066	0.0103	0.0055	72	-2.10	027.43	0
027	Magnesium (%)	0407	0.1873	0.0178	0.2066	0.0103	0.0055	72	-1.87	027.41	0
027	Magnesium (%)	0083	0.1900	0.0005	0.2066	0.0103	0.0055	72	-1.62	027.43	0
027	Magnesium (%)	0089	0.1900	0.0000	0.2066	0.0103	0.0055	72	-1.61	027.43	0
027	Magnesium (%)	2302	0.1900	0.0000	0.2066	0.0103	0.0055	72	-1.61	027.99	0
027	Magnesium (%)	0407	0.1920	0.0079	0.2066	0.0103	0.0055	72	-1.42	027.43	0
027	Magnesium (%)	0038	0.1930	0.0160	0.2066	0.0103	0.0055	72	-1.32	027.43	0
027	Magnesium (%)	2129	0.1932	0.0044	0.2066	0.0103	0.0055	72	-1.30	027.42	0
027	Magnesium (%)	0186	0.1948	0.0024	0.2066	0.0103	0.0055	72	-1.15	027.52	0
027	Magnesium (%)	0038	0.1959	0.0007	0.2066	0.0103	0.0055	72	-1.04	027.31	0
027	Magnesium (%)	0560	0.1986	0.0108	0.2066	0.0103	0.0055	72	-0.78	027.52	0
027	Magnesium (%)	0154	0.1988	0.0074	0.2066	0.0103	0.0055	72	-0.76	027.52	0
027	Magnesium (%)	0202	0.1995	0.0030	0.2066	0.0103	0.0055	72	-0.69	027.43	0
027	Magnesium (%)	0229	0.2000	0.0000	0.2066	0.0103	0.0055	72	-0.64	027.41	0
027	Magnesium (%)	0520	0.2000	0.0000	0.2066	0.0103	0.0055	72	-0.64	027.41	0
027	Magnesium (%)	0910	0.2000	0.0000	0.2066	0.0103	0.0055	72	-0.64	027.41	0
027	Magnesium (%)	0294	0.2000	0.0000	0.2066	0.0103	0.0055	72	-0.64	027.42	0
027	Magnesium (%)	0692	0.2000	0.0200	0.2066	0.0103	0.0055	72	-0.64	027.42	0
027	Magnesium (%)	0035	0.2000	0.0000	0.2066	0.0103	0.0055	72	-0.64	027.43	0
027	Magnesium (%)	0675	0.2000	0.0000	0.2066	0.0103	0.0055	72	-0.64	027.43	0
027	Magnesium (%)	2033	0.2000	0.0000	0.2066	0.0103	0.0055	72	-0.64	027.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
027	Magnesium (%)	0242	0.2000	0.0000	0.2066	0.0103	0.0055	72	-0.64	027.99	0
027	Magnesium (%)	0870	0.2014	0.0064	0.2066	0.0103	0.0055	72	-0.51	027.42	0
027	Magnesium (%)	0345	0.2025	0.0010	0.2066	0.0103	0.0055	72	-0.40	027.43	0
027	Magnesium (%)	0357	0.2029	0.0049	0.2066	0.0103	0.0055	72	-0.37	027.42	0
027	Magnesium (%)	0968	0.2035	0.0010	0.2066	0.0103	0.0055	72	-0.30	027.43	0
027	Magnesium (%)	0939	0.2050	0.0100	0.2066	0.0103	0.0055	72	-0.16	027.31	0
027	Magnesium (%)	0169	0.2050	0.0100	0.2066	0.0103	0.0055	72	-0.16	027.32	0
027	Magnesium (%)	0003	0.2050	0.0100	0.2066	0.0103	0.0055	72	-0.16	027.41	0
027	Magnesium (%)	0878	0.2050	0.0100	0.2066	0.0103	0.0055	72	-0.16	027.41	0
027	Magnesium (%)	0265	0.2050	0.0100	0.2066	0.0103	0.0055	72	-0.16	027.42	0
027	Magnesium (%)	0278	0.2050	0.0100	0.2066	0.0103	0.0055	72	-0.16	027.42	0
027	Magnesium (%)	0297	0.2050	0.0100	0.2066	0.0103	0.0055	72	-0.16	027.43	0
027	Magnesium (%)	0939	0.2050	0.0100	0.2066	0.0103	0.0055	72	-0.16	027.53	0
027	Magnesium (%)	0590	0.2050	0.0100	0.2066	0.0103	0.0055	72	-0.16	027.99	0
027	Magnesium (%)	0142	0.2066	0.0044	0.2066	0.0103	0.0055	72	0.00	027.31	0
027	Magnesium (%)	0263	0.2070	0.0020	0.2066	0.0103	0.0055	72	0.04	027.42	0
027	Magnesium (%)	0164	0.2070	0.0100	0.2066	0.0103	0.0055	72	0.04	027.44	0
027	Magnesium (%)	0074	0.2090	0.0000	0.2066	0.0103	0.0055	72	0.23	027.41	0
027	Magnesium (%)	0148	0.2090	0.0020	0.2066	0.0103	0.0055	72	0.23	027.41	0
027	Magnesium (%)	0868	0.2100	0.0000	0.2066	0.0103	0.0055	72	0.33	027.31	0
027	Magnesium (%)	2196	0.2100	0.0000	0.2066	0.0103	0.0055	72	0.33	027.31	0
027	Magnesium (%)	2012	0.2100	0.0000	0.2066	0.0103	0.0055	72	0.33	027.41	0
027	Magnesium (%)	0035	0.2100	0.0000	0.2066	0.0103	0.0055	72	0.33	027.42	0
027	Magnesium (%)	0726	0.2100	0.0000	0.2066	0.0103	0.0055	72	0.33	027.42	0
027	Magnesium (%)	0510	0.2100	0.0000	0.2066	0.0103	0.0055	72	0.33	027.43	0
027	Magnesium (%)	0918	0.2100	0.0020	0.2066	0.0103	0.0055	72	0.33	027.43	0
027	Magnesium (%)	0948	0.2100	0.0200	0.2066	0.0103	0.0055	72	0.33	027.33	0
027	Magnesium (%)	0358	0.2100	0.0200	0.2066	0.0103	0.0055	72	0.33	027.41	0
027	Magnesium (%)	0553	0.2105	0.0110	0.2066	0.0103	0.0055	72	0.37	027.53	0
027	Magnesium (%)	0870	0.2112	0.0068	0.2066	0.0103	0.0055	72	0.44	027.43	0
027	Magnesium (%)	0045	0.2115	0.0050	0.2066	0.0103	0.0055	72	0.47	027.42	0
027	Magnesium (%)	0171	0.2130	0.0020	0.2066	0.0103	0.0055	72	0.62	027.41	0
027	Magnesium (%)	0572	0.2135	0.0090	0.2066	0.0103	0.0055	72	0.67	027.53	0
027	Magnesium (%)	0536	0.2150	0.0100	0.2066	0.0103	0.0055	72	0.81	027.31	0
027	Magnesium (%)	0689	0.2150	0.0100	0.2066	0.0103	0.0055	72	0.81	027.31	0
027	Magnesium (%)	0190	0.2150	0.0100	0.2066	0.0103	0.0055	72	0.81	027.42	0
027	Magnesium (%)	0425	0.2150	0.0100	0.2066	0.0103	0.0055	72	0.81	027.43	0
027	Magnesium (%)	0563	0.2158	0.0014	0.2066	0.0103	0.0055	72	0.89	027.31	0
027	Magnesium (%)	0693	0.2180	0.0040	0.2066	0.0103	0.0055	72	1.10	027.42	0
027	Magnesium (%)	0964	0.2189	0.0020	0.2066	0.0103	0.0055	72	1.19	027.43	0
027	Magnesium (%)	0098	0.2200	0.0000	0.2066	0.0103	0.0055	72	1.30	027.41	0
027	Magnesium (%)	0226	0.2200	0.0000	0.2066	0.0103	0.0055	72	1.30	027.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
027	Magnesium (%)	0037	0.2200	0.0000	0.2066	0.0103	0.0055	72	1.30	027.42	0
027	Magnesium (%)	0682	0.2200	0.0000	0.2066	0.0103	0.0055	72	1.30	027.43	0
027	Magnesium (%)	0100	0.2200	0.0000	0.2066	0.0103	0.0055	72	1.30	027.99	0
027	Magnesium (%)	0186	0.2210	0.0102	0.2066	0.0103	0.0055	72	1.39	027.42	0
027	Magnesium (%)	0598	0.2213	0.0009	0.2066	0.0103	0.0055	72	1.42	027.41	0
027	Magnesium (%)	0354	0.2250	0.0100	0.2066	0.0103	0.0055	72	1.78	027.41	0
027	Magnesium (%)	0098	0.2250	0.0100	0.2066	0.0103	0.0055	72	1.78	027.43	0
027	Magnesium (%)	0042	0.2335	0.0110	0.2066	0.0103	0.0055	72	2.60	027.42	0
028	Manganese (ppm)	0921	102.5	1.000	123.2	7.881	3.307	86	-2.63	028.31	0
028	Manganese (ppm)	0354	103.9	1.220	123.2	7.881	3.307	86	-2.45	028.41	0
028	Manganese (ppm)	0536	104.3	2.000	123.2	7.881	3.307	86	-2.40	028.31	0
028	Manganese (ppm)	0619	104.7	0.7150	123.2	7.881	3.307	86	-2.35	028.41	0
028	Manganese (ppm)	0047	105.1	0.7000	123.2	7.881	3.307	86	-2.30	028.52	0
028	Manganese (ppm)	0870	106.0	7.200	123.2	7.881	3.307	86	-2.18	028.42	0
028	Manganese (ppm)	0870	107.9	4.000	123.2	7.881	3.307	86	-1.94	028.43	0
028	Manganese (ppm)	2129	110.2	0.2000	123.2	7.881	3.307	86	-1.65	028.42	0
028	Manganese (ppm)	0939	110.3	1.740	123.2	7.881	3.307	86	-1.64	028.31	0
028	Manganese (ppm)	0910	111.0	8.000	123.2	7.881	3.307	86	-1.55	028.41	0
028	Manganese (ppm)	0226	112.7	5.010	123.2	7.881	3.307	86	-1.33	028.41	0
028	Manganese (ppm)	2146	113.5	4.170	123.2	7.881	3.307	86	-1.23	028.31	0
028	Manganese (ppm)	0066	113.6	4.236	123.2	7.881	3.307	86	-1.22	028.44	0
028	Manganese (ppm)	0171	113.9	0.1000	123.2	7.881	3.307	86	-1.18	028.41	0
028	Manganese (ppm)	0505	116.9	1.100	123.2	7.881	3.307	86	-0.80	028.43	0
028	Manganese (ppm)	0001	117.6	0.3000	123.2	7.881	3.307	86	-0.72	028.31	0
028	Manganese (ppm)	2012	117.7	3.120	123.2	7.881	3.307	86	-0.70	028.41	0
028	Manganese (ppm)	0948	118.0	0.5700	123.2	7.881	3.307	86	-0.66	028.33	0
028	Manganese (ppm)	2302	118.0	0.0000	123.2	7.881	3.307	86	-0.66	028.99	0
028	Manganese (ppm)	0425	118.6	7.700	123.2	7.881	3.307	86	-0.59	028.43	0
028	Manganese (ppm)	0278	119.0	2.000	123.2	7.881	3.307	86	-0.53	028.42	0
028	Manganese (ppm)	2196	119.2	0.0000	123.2	7.881	3.307	86	-0.51	028.31	0
028	Manganese (ppm)	0878	119.5	1.000	123.2	7.881	3.307	86	-0.47	028.41	0
028	Manganese (ppm)	0553	119.5	9.000	123.2	7.881	3.307	86	-0.47	028.53	0
028	Manganese (ppm)	0045	120.0	12.00	123.2	7.881	3.307	86	-0.40	028.42	0
028	Manganese (ppm)	0297	120.0	4.000	123.2	7.881	3.307	86	-0.40	028.43	0
028	Manganese (ppm)	0689	120.3	1.400	123.2	7.881	3.307	86	-0.37	028.31	0
028	Manganese (ppm)	0357	120.6	1.618	123.2	7.881	3.307	86	-0.33	028.42	0
028	Manganese (ppm)	0693	120.9	4.520	123.2	7.881	3.307	86	-0.30	028.42	0
028	Manganese (ppm)	0169	121.5	1.000	123.2	7.881	3.307	86	-0.21	028.43	0
028	Manganese (ppm)	0242	121.5	7.000	123.2	7.881	3.307	86	-0.21	028.99	0
028	Manganese (ppm)	0918	121.5	4.170	123.2	7.881	3.307	86	-0.21	028.43	0
028	Manganese (ppm)	0229	121.7	2.740	123.2	7.881	3.307	86	-0.19	028.41	0
028	Manganese (ppm)	0407	121.7	8.729	123.2	7.881	3.307	86	-0.19	028.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
028	Manganese (ppm)	0572	122.0	6.000	123.2	7.881	3.307	86	-0.15	028.53	0
028	Manganese (ppm)	0675	122.1	1.550	123.2	7.881	3.307	86	-0.14	028.43	0
028	Manganese (ppm)	0148	122.1	3.536	123.2	7.881	3.307	86	-0.13	028.41	0
028	Manganese (ppm)	2192	122.3	2.010	123.2	7.881	3.307	86	-0.11	028.53	0
028	Manganese (ppm)	0263	122.4	0.5290	123.2	7.881	3.307	86	-0.10	028.42	0
028	Manganese (ppm)	0407	122.5	0.8725	123.2	7.881	3.307	86	-0.09	028.43	0
028	Manganese (ppm)	0009	122.5	5.000	123.2	7.881	3.307	86	-0.09	028.42	0
028	Manganese (ppm)	0563	122.5	0.9992	123.2	7.881	3.307	86	-0.09	028.31	0
028	Manganese (ppm)	0904	122.8	14.80	123.2	7.881	3.307	86	-0.05	028.33	0
028	Manganese (ppm)	0019	123.0	2.500	123.2	7.881	3.307	86	-0.03	028.41	0
028	Manganese (ppm)	0074	123.0	2.000	123.2	7.881	3.307	86	-0.02	028.41	0
028	Manganese (ppm)	0590	123.0	12.00	123.2	7.881	3.307	86	-0.02	028.99	0
028	Manganese (ppm)	0038	123.5	1.000	123.2	7.881	3.307	86	0.04	028.43	0
028	Manganese (ppm)	0510	123.5	1.000	123.2	7.881	3.307	86	0.04	028.43	0
028	Manganese (ppm)	0560	124.2	0.5000	123.2	7.881	3.307	86	0.12	028.52	0
028	Manganese (ppm)	2033	124.2	1.150	123.2	7.881	3.307	86	0.13	028.43	0
028	Manganese (ppm)	0692	124.5	1.000	123.2	7.881	3.307	86	0.17	028.42	0
028	Manganese (ppm)	0083	124.5	1.000	123.2	7.881	3.307	86	0.17	028.43	0
028	Manganese (ppm)	0353	124.5	1.000	123.2	7.881	3.307	86	0.17	028.43	0
028	Manganese (ppm)	0164	124.5	1.000	123.2	7.881	3.307	86	0.17	028.44	0
028	Manganese (ppm)	0175	125.0	2.000	123.2	7.881	3.307	86	0.23	028.31	0
028	Manganese (ppm)	0939	125.0	2.000	123.2	7.881	3.307	86	0.23	028.53	0
028	Manganese (ppm)	0529	125.3	2.600	123.2	7.881	3.307	86	0.27	028.31	0
028	Manganese (ppm)	0520	125.5	9.000	123.2	7.881	3.307	86	0.29	028.41	0
028	Manganese (ppm)	0265	126.0	2.000	123.2	7.881	3.307	86	0.36	028.42	0
028	Manganese (ppm)	0358	126.2	11.48	123.2	7.881	3.307	86	0.38	028.41	0
028	Manganese (ppm)	0345	126.5	3.000	123.2	7.881	3.307	86	0.42	028.43	0
028	Manganese (ppm)	0914	126.8	0.8000	123.2	7.881	3.307	86	0.46	028.43	0
028	Manganese (ppm)	0186	128.0	2.000	123.2	7.881	3.307	86	0.61	028.42	0
028	Manganese (ppm)	0100	128.0	6.000	123.2	7.881	3.307	86	0.61	028.99	0
028	Manganese (ppm)	0004	128.0	7.250	123.2	7.881	3.307	86	0.62	028.41	0
028	Manganese (ppm)	0868	128.5	3.000	123.2	7.881	3.307	86	0.67	028.31	0
028	Manganese (ppm)	0968	128.5	1.000	123.2	7.881	3.307	86	0.67	028.43	0
028	Manganese (ppm)	0656	129.5	0.0000	123.2	7.881	3.307	86	0.80	028.31	0
028	Manganese (ppm)	0035	129.5	1.000	123.2	7.881	3.307	86	0.80	028.42	0
028	Manganese (ppm)	0598	129.6	6.600	123.2	7.881	3.307	86	0.81	028.41	0
028	Manganese (ppm)	0294	130.1	0.6000	123.2	7.881	3.307	86	0.88	028.42	0
028	Manganese (ppm)	0964	130.5	11.35	123.2	7.881	3.307	86	0.92	028.43	0
028	Manganese (ppm)	0003	130.5	3.000	123.2	7.881	3.307	86	0.93	028.41	0
028	Manganese (ppm)	0682	130.7	0.0000	123.2	7.881	3.307	86	0.96	028.43	0
028	Manganese (ppm)	2234	131.9	1.670	123.2	7.881	3.307	86	1.11	028.43	0
028	Manganese (ppm)	0186	133.0	4.000	123.2	7.881	3.307	86	1.25	028.52	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
028	Manganese (ppm)	0098	133.9	8.500	123.2	7.881	3.307	86	1.35	028.43	0
028	Manganese (ppm)	0042	134.0	8.000	123.2	7.881	3.307	86	1.37	028.42	0
028	Manganese (ppm)	0098	134.2	0.1000	123.2	7.881	3.307	86	1.39	028.41	0
028	Manganese (ppm)	0726	136.1	0.4000	123.2	7.881	3.307	86	1.64	028.42	0
028	Manganese (ppm)	0190	137.7	0.7000	123.2	7.881	3.307	86	1.84	028.42	0
028	Manganese (ppm)	0202	140.5	2.400	123.2	7.881	3.307	86	2.20	028.43	0
028	Manganese (ppm)	0037	142.7	1.200	123.2	7.881	3.307	86	2.48	028.42	0
028	Manganese (ppm)	0035	144.0	2.000	123.2	7.881	3.307	86	2.64	028.43	0
028	Manganese (ppm)	2188	144.8	0.0700	123.2	7.881	3.307	86	2.75	028.31	0
028	Manganese (ppm)	0038	146.5	9.000	123.2	7.881	3.307	86	2.96	028.32	0
031	Phosphorus (%)	0186	0.6153	0.0077	0.7597	0.0369	0.0189	117	-3.91	031.42	0
031	Phosphorus (%)	2302	0.6600	0.0000	0.7597	0.0369	0.0189	117	-2.70	031.01	0
031	Phosphorus (%)	0278	0.6700	0.0400	0.7597	0.0369	0.0189	117	-2.43	031.42	0
031	Phosphorus (%)	0186	0.6798	0.0223	0.7597	0.0369	0.0189	117	-2.16	031.52	0
031	Phosphorus (%)	0047	0.6850	0.0500	0.7597	0.0369	0.0189	117	-2.02	031.42	0
031	Phosphorus (%)	2302	0.6900	0.0000	0.7597	0.0369	0.0189	117	-1.89	031.99	0
031	Phosphorus (%)	0848	0.6950	0.0100	0.7597	0.0369	0.0189	117	-1.75	031.41	0
031	Phosphorus (%)	0693	0.6990	0.0100	0.7597	0.0369	0.0189	117	-1.64	031.42	0
031	Phosphorus (%)	2144	0.7000	0.0000	0.7597	0.0369	0.0189	117	-1.62	031.01	0
031	Phosphorus (%)	2033	0.7050	0.0040	0.7597	0.0369	0.0189	117	-1.48	031.43	0
031	Phosphorus (%)	0870	0.7079	0.0384	0.7597	0.0369	0.0189	117	-1.40	031.42	0
031	Phosphorus (%)	0358	0.7100	0.0200	0.7597	0.0369	0.0189	117	-1.34	031.41	0
031	Phosphorus (%)	0242	0.7150	0.0300	0.7597	0.0369	0.0189	117	-1.21	031.99	0
031	Phosphorus (%)	0560	0.7151	0.0827	0.7597	0.0369	0.0189	117	-1.21	031.52	0
031	Phosphorus (%)	0626	0.7200	0.0600	0.7597	0.0369	0.0189	117	-1.07	031.01	0
031	Phosphorus (%)	2146	0.7200	0.0000	0.7597	0.0369	0.0189	117	-1.07	031.01	0
031	Phosphorus (%)	0910	0.7200	0.0000	0.7597	0.0369	0.0189	117	-1.07	031.41	0
031	Phosphorus (%)	0265	0.7200	0.0400	0.7597	0.0369	0.0189	117	-1.07	031.42	0
031	Phosphorus (%)	2129	0.7224	0.0058	0.7597	0.0369	0.0189	117	-1.01	031.42	0
031	Phosphorus (%)	2022	0.7250	0.0100	0.7597	0.0369	0.0189	117	-0.94	031.01	0
031	Phosphorus (%)	0536	0.7250	0.0100	0.7597	0.0369	0.0189	117	-0.94	031.06	0
031	Phosphorus (%)	0144	0.7250	0.0900	0.7597	0.0369	0.0189	117	-0.94	031.41	0
031	Phosphorus (%)	2259	0.7270	0.0620	0.7597	0.0369	0.0189	117	-0.88	031.41	0
031	Phosphorus (%)	0037	0.7285	0.0130	0.7597	0.0369	0.0189	117	-0.84	031.42	0
031	Phosphorus (%)	0939	0.7300	0.0200	0.7597	0.0369	0.0189	117	-0.80	031.01	0
031	Phosphorus (%)	0009	0.7300	0.0800	0.7597	0.0369	0.0189	117	-0.80	031.42	0
031	Phosphorus (%)	0692	0.7300	0.0400	0.7597	0.0369	0.0189	117	-0.80	031.42	0
031	Phosphorus (%)	0918	0.7345	0.0050	0.7597	0.0369	0.0189	117	-0.68	031.43	0
031	Phosphorus (%)	0019	0.7350	0.0100	0.7597	0.0369	0.0189	117	-0.67	031.41	0
031	Phosphorus (%)	2246	0.7400	0.0000	0.7597	0.0369	0.0189	117	-0.53	031.01	0
031	Phosphorus (%)	0676	0.7400	0.0000	0.7597	0.0369	0.0189	117	-0.53	031.99	0
031	Phosphorus (%)	0553	0.7420	0.0120	0.7597	0.0369	0.0189	117	-0.48	031.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
031	Phosphorus (%)	0175	0.7450	0.0100	0.7597	0.0369	0.0189	117	-0.40	031.01	0
031	Phosphorus (%)	0294	0.7450	0.0300	0.7597	0.0369	0.0189	117	-0.40	031.42	0
031	Phosphorus (%)	0297	0.7450	0.0500	0.7597	0.0369	0.0189	117	-0.40	031.43	0
031	Phosphorus (%)	2109	0.7465	0.0010	0.7597	0.0369	0.0189	117	-0.36	031.01	0
031	Phosphorus (%)	0407	0.7474	0.0832	0.7597	0.0369	0.0189	117	-0.33	031.41	0
031	Phosphorus (%)	0407	0.7480	0.0046	0.7597	0.0369	0.0189	117	-0.32	031.43	0
031	Phosphorus (%)	2062	0.7491	0.0052	0.7597	0.0369	0.0189	117	-0.29	031.01	0
031	Phosphorus (%)	0893	0.7500	0.0000	0.7597	0.0369	0.0189	117	-0.26	031.00	0
031	Phosphorus (%)	0529	0.7500	0.0200	0.7597	0.0369	0.0189	117	-0.26	031.01	0
031	Phosphorus (%)	0590	0.7500	0.0400	0.7597	0.0369	0.0189	117	-0.26	031.01	0
031	Phosphorus (%)	0674	0.7500	0.0000	0.7597	0.0369	0.0189	117	-0.26	031.01	0
031	Phosphorus (%)	0298	0.7500	0.0200	0.7597	0.0369	0.0189	117	-0.26	031.41	0
031	Phosphorus (%)	0354	0.7500	0.0200	0.7597	0.0369	0.0189	117	-0.26	031.41	0
031	Phosphorus (%)	0089	0.7500	0.0000	0.7597	0.0369	0.0189	117	-0.26	031.43	0
031	Phosphorus (%)	2234	0.7500	0.0600	0.7597	0.0369	0.0189	117	-0.26	031.43	0
031	Phosphorus (%)	0100	0.7500	0.0200	0.7597	0.0369	0.0189	117	-0.26	031.99	0
031	Phosphorus (%)	2009	0.7501	0.0002	0.7597	0.0369	0.0189	117	-0.26	031.01	0
031	Phosphorus (%)	0036	0.7510	0.0064	0.7597	0.0369	0.0189	117	-0.23	031.03	0
031	Phosphorus (%)	0001	0.7548	0.0082	0.7597	0.0369	0.0189	117	-0.13	031.03	0
031	Phosphorus (%)	0904	0.7550	0.0300	0.7597	0.0369	0.0189	117	-0.13	031.01	0
031	Phosphorus (%)	2196	0.7550	0.0000	0.7597	0.0369	0.0189	117	-0.13	031.01	0
031	Phosphorus (%)	0045	0.7550	0.0760	0.7597	0.0369	0.0189	117	-0.13	031.42	0
031	Phosphorus (%)	0675	0.7550	0.0100	0.7597	0.0369	0.0189	117	-0.13	031.43	0
031	Phosphorus (%)	0939	0.7550	0.0100	0.7597	0.0369	0.0189	117	-0.13	031.53	0
031	Phosphorus (%)	0345	0.7560	0.0080	0.7597	0.0369	0.0189	117	-0.10	031.43	0
031	Phosphorus (%)	0042	0.7595	0.0010	0.7597	0.0369	0.0189	117	0.00	031.42	0
031	Phosphorus (%)	0656	0.7600	0.0000	0.7597	0.0369	0.0189	117	0.01	031.01	0
031	Phosphorus (%)	0683	0.7600	0.0000	0.7597	0.0369	0.0189	117	0.01	031.01	0
031	Phosphorus (%)	0013	0.7600	0.0200	0.7597	0.0369	0.0189	117	0.01	031.02	0
031	Phosphorus (%)	0425	0.7600	0.0400	0.7597	0.0369	0.0189	117	0.01	031.43	0
031	Phosphorus (%)	0036	0.7626	0.0200	0.7597	0.0369	0.0189	117	0.08	031.43	0
031	Phosphorus (%)	0870	0.7646	0.0313	0.7597	0.0369	0.0189	117	0.13	031.43	0
031	Phosphorus (%)	2188	0.7650	0.0100	0.7597	0.0369	0.0189	117	0.14	031.01	0
031	Phosphorus (%)	0169	0.7650	0.0100	0.7597	0.0369	0.0189	117	0.14	031.03	0
031	Phosphorus (%)	0004	0.7650	0.0300	0.7597	0.0369	0.0189	117	0.14	031.41	0
031	Phosphorus (%)	0123	0.7650	0.0100	0.7597	0.0369	0.0189	117	0.14	031.41	0
031	Phosphorus (%)	0171	0.7650	0.0100	0.7597	0.0369	0.0189	117	0.14	031.41	0
031	Phosphorus (%)	0226	0.7650	0.0100	0.7597	0.0369	0.0189	117	0.14	031.41	0
031	Phosphorus (%)	0142	0.7655	0.0173	0.7597	0.0369	0.0189	117	0.16	031.01	0
031	Phosphorus (%)	0881	0.7660	0.0020	0.7597	0.0369	0.0189	117	0.17	031.01	0
031	Phosphorus (%)	0948	0.7680	0.0020	0.7597	0.0369	0.0189	117	0.23	031.01	0
031	Phosphorus (%)	2006	0.7680	0.0060	0.7597	0.0369	0.0189	117	0.23	031.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	0164	0.7680	0.0020	0.7597	0.0369	0.0189	117	0.23	031.44	0
031	Phosphorus (%)	0689	0.7700	0.0000	0.7597	0.0369	0.0189	117	0.28	031.01	0
031	Phosphorus (%)	0868	0.7700	0.0000	0.7597	0.0369	0.0189	117	0.28	031.01	0
031	Phosphorus (%)	0035	0.7700	0.0000	0.7597	0.0369	0.0189	117	0.28	031.42	0
031	Phosphorus (%)	0083	0.7700	0.0000	0.7597	0.0369	0.0189	117	0.28	031.43	0
031	Phosphorus (%)	0353	0.7700	0.0400	0.7597	0.0369	0.0189	117	0.28	031.43	0
031	Phosphorus (%)	2152	0.7700	0.0200	0.7597	0.0369	0.0189	117	0.28	031.44	0
031	Phosphorus (%)	0357	0.7743	0.0148	0.7597	0.0369	0.0189	117	0.40	031.42	0
031	Phosphorus (%)	0229	0.7750	0.0100	0.7597	0.0369	0.0189	117	0.41	031.41	0
031	Phosphorus (%)	0010	0.7750	0.0300	0.7597	0.0369	0.0189	117	0.41	031.43	0
031	Phosphorus (%)	0968	0.7750	0.0080	0.7597	0.0369	0.0189	117	0.41	031.43	0
031	Phosphorus (%)	0878	0.7800	0.0200	0.7597	0.0369	0.0189	117	0.55	031.01	0
031	Phosphorus (%)	0897	0.7800	0.0000	0.7597	0.0369	0.0189	117	0.55	031.01	0
031	Phosphorus (%)	0682	0.7800	0.0000	0.7597	0.0369	0.0189	117	0.55	031.43	0
031	Phosphorus (%)	2295	0.7850	0.0100	0.7597	0.0369	0.0189	117	0.69	031.01	0
031	Phosphorus (%)	2012	0.7850	0.0100	0.7597	0.0369	0.0189	117	0.69	031.41	0
031	Phosphorus (%)	0885	0.7866	0.0008	0.7597	0.0369	0.0189	117	0.73	031.01	0
031	Phosphorus (%)	0038	0.7875	0.0070	0.7597	0.0369	0.0189	117	0.75	031.43	0
031	Phosphorus (%)	0263	0.7905	0.0030	0.7597	0.0369	0.0189	117	0.83	031.42	0
031	Phosphorus (%)	0202	0.7910	0.0040	0.7597	0.0369	0.0189	117	0.85	031.43	0
031	Phosphorus (%)	0563	0.7912	0.0031	0.7597	0.0369	0.0189	117	0.85	031.01	0
031	Phosphorus (%)	0572	0.7915	0.0510	0.7597	0.0369	0.0189	117	0.86	031.53	0
031	Phosphorus (%)	0148	0.7925	0.0050	0.7597	0.0369	0.0189	117	0.89	031.41	0
031	Phosphorus (%)	0035	0.7950	0.0100	0.7597	0.0369	0.0189	117	0.96	031.43	0
031	Phosphorus (%)	0687	0.8000	0.1000	0.7597	0.0369	0.0189	117	1.09	031.01	0
031	Phosphorus (%)	0505	0.8000	0.0400	0.7597	0.0369	0.0189	117	1.09	031.02	0
031	Phosphorus (%)	0066	0.8035	0.0310	0.7597	0.0369	0.0189	117	1.19	031.44	0
031	Phosphorus (%)	0190	0.8050	0.0100	0.7597	0.0369	0.0189	117	1.23	031.42	0
031	Phosphorus (%)	0726	0.8050	0.0100	0.7597	0.0369	0.0189	117	1.23	031.42	0
031	Phosphorus (%)	0154	0.8067	0.0091	0.7597	0.0369	0.0189	117	1.27	031.52	0
031	Phosphorus (%)	0001	0.8100	0.0100	0.7597	0.0369	0.0189	117	1.36	031.43	0
031	Phosphorus (%)	0098	0.8104	0.0220	0.7597	0.0369	0.0189	117	1.37	031.41	0
031	Phosphorus (%)	2053	0.8150	0.0500	0.7597	0.0369	0.0189	117	1.50	031.02	0
031	Phosphorus (%)	0074	0.8185	0.0270	0.7597	0.0369	0.0189	117	1.59	031.41	0
031	Phosphorus (%)	0914	0.8239	0.0686	0.7597	0.0369	0.0189	117	1.74	031.43	0
031	Phosphorus (%)	0520	0.8250	0.0100	0.7597	0.0369	0.0189	117	1.77	031.41	0
031	Phosphorus (%)	0510	0.8250	0.0100	0.7597	0.0369	0.0189	117	1.77	031.43	0
031	Phosphorus (%)	0964	0.8439	0.0203	0.7597	0.0369	0.0189	117	2.28	031.43	0
031	Phosphorus (%)	0098	0.8452	0.0023	0.7597	0.0369	0.0189	117	2.31	031.43	0
031	Phosphorus (%)	0598	0.8549	0.0073	0.7597	0.0369	0.0189	117	2.58	031.41	0
031	Phosphorus (%)	0619	0.8560	0.0060	0.7597	0.0369	0.0189	117	2.61	031.41	0
031	Phosphorus (%)	2192	0.9250	0.0100	0.7597	0.0369	0.0189	117	4.47	031.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	2181	1.165	0.0100	0.7597	0.0369	0.0189	117	10.97	031.01	0
031	Phosphorus (%)	0003	1.120	0.5400	0.7597	0.0369	0.0189	117	9.75	031.41	1
031	Phosphorus (%)	0108	2.790	0.0600	0.7597	0.0369	0.0189	117	54.95	031.01	2
032	Potassium (%)	0019	0.6550	0.0300	0.8001	0.0411	0.0194	73	-3.53	032.41	0
032	Potassium (%)	0358	0.7200	0.0200	0.8001	0.0411	0.0194	73	-1.95	032.41	0
032	Potassium (%)	0563	0.7329	0.0018	0.8001	0.0411	0.0194	73	-1.63	032.31	0
032	Potassium (%)	0144	0.7400	0.0000	0.8001	0.0411	0.0194	73	-1.46	032.41	0
032	Potassium (%)	2012	0.7400	0.0200	0.8001	0.0411	0.0194	73	-1.46	032.41	0
032	Potassium (%)	0003	0.7450	0.0100	0.8001	0.0411	0.0194	73	-1.34	032.41	0
032	Potassium (%)	0425	0.7450	0.0700	0.8001	0.0411	0.0194	73	-1.34	032.43	0
032	Potassium (%)	0675	0.7450	0.0100	0.8001	0.0411	0.0194	73	-1.34	032.43	0
032	Potassium (%)	0910	0.7500	0.0600	0.8001	0.0411	0.0194	73	-1.22	032.41	0
032	Potassium (%)	0354	0.7550	0.0100	0.8001	0.0411	0.0194	73	-1.10	032.41	0
032	Potassium (%)	0278	0.7550	0.0100	0.8001	0.0411	0.0194	73	-1.10	032.42	0
032	Potassium (%)	0918	0.7590	0.0300	0.8001	0.0411	0.0194	73	-1.00	032.43	0
032	Potassium (%)	0870	0.7640	0.0380	0.8001	0.0411	0.0194	73	-0.88	032.42	0
032	Potassium (%)	0294	0.7650	0.0100	0.8001	0.0411	0.0194	73	-0.85	032.42	0
032	Potassium (%)	0590	0.7650	0.0300	0.8001	0.0411	0.0194	73	-0.85	032.99	0
032	Potassium (%)	0939	0.7700	0.0000	0.8001	0.0411	0.0194	73	-0.73	032.31	0
032	Potassium (%)	0553	0.7730	0.0480	0.8001	0.0411	0.0194	73	-0.66	032.53	0
032	Potassium (%)	0229	0.7750	0.0100	0.8001	0.0411	0.0194	73	-0.61	032.41	0
032	Potassium (%)	0142	0.7771	0.0141	0.8001	0.0411	0.0194	73	-0.56	032.31	0
032	Potassium (%)	0154	0.7785	0.0130	0.8001	0.0411	0.0194	73	-0.52	032.52	0
032	Potassium (%)	0683	0.7800	0.0000	0.8001	0.0411	0.0194	73	-0.49	032.31	0
032	Potassium (%)	0265	0.7800	0.0600	0.8001	0.0411	0.0194	73	-0.49	032.42	0
032	Potassium (%)	2033	0.7800	0.0200	0.8001	0.0411	0.0194	73	-0.49	032.43	0
032	Potassium (%)	0100	0.7800	0.0200	0.8001	0.0411	0.0194	73	-0.49	032.99	0
032	Potassium (%)	2302	0.7800	0.0000	0.8001	0.0411	0.0194	73	-0.49	032.99	0
032	Potassium (%)	0560	0.7847	0.0753	0.8001	0.0411	0.0194	73	-0.38	032.52	0
032	Potassium (%)	0536	0.7850	0.0100	0.8001	0.0411	0.0194	73	-0.37	032.31	0
032	Potassium (%)	0868	0.7850	0.0100	0.8001	0.0411	0.0194	73	-0.37	032.31	0
032	Potassium (%)	0083	0.7850	0.0100	0.8001	0.0411	0.0194	73	-0.37	032.43	0
032	Potassium (%)	0357	0.7873	0.0179	0.8001	0.0411	0.0194	73	-0.31	032.42	0
032	Potassium (%)	0038	0.7880	0.0420	0.8001	0.0411	0.0194	73	-0.29	032.43	0
032	Potassium (%)	0186	0.7886	0.0207	0.8001	0.0411	0.0194	73	-0.28	032.52	0
032	Potassium (%)	2129	0.7930	0.0010	0.8001	0.0411	0.0194	73	-0.17	032.42	0
032	Potassium (%)	2053	0.7950	0.0500	0.8001	0.0411	0.0194	73	-0.12	032.31	0
032	Potassium (%)	0870	0.7991	0.0137	0.8001	0.0411	0.0194	73	-0.02	032.43	0
032	Potassium (%)	0682	0.8000	0.0000	0.8001	0.0411	0.0194	73	0.00	032.43	0
032	Potassium (%)	0242	0.8000	0.0200	0.8001	0.0411	0.0194	73	0.00	032.99	0
032	Potassium (%)	0186	0.8005	0.0140	0.8001	0.0411	0.0194	73	0.01	032.42	0
032	Potassium (%)	0693	0.8010	0.0080	0.8001	0.0411	0.0194	73	0.02	032.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
032	Potassium (%)	0042	0.8015	0.0030	0.8001	0.0411	0.0194	73	0.03	032.42	0
032	Potassium (%)	0345	0.8040	0.0020	0.8001	0.0411	0.0194	73	0.10	032.43	0
032	Potassium (%)	0169	0.8050	0.0100	0.8001	0.0411	0.0194	73	0.12	032.32	0
032	Potassium (%)	0878	0.8050	0.0100	0.8001	0.0411	0.0194	73	0.12	032.41	0
032	Potassium (%)	0164	0.8105	0.0010	0.8001	0.0411	0.0194	73	0.25	032.44	0
032	Potassium (%)	0074	0.8120	0.0180	0.8001	0.0411	0.0194	73	0.29	032.41	0
032	Potassium (%)	0035	0.8150	0.0100	0.8001	0.0411	0.0194	73	0.36	032.42	0
032	Potassium (%)	0035	0.8150	0.0100	0.8001	0.0411	0.0194	73	0.36	032.43	0
032	Potassium (%)	0353	0.8150	0.0100	0.8001	0.0411	0.0194	73	0.36	032.43	0
032	Potassium (%)	0037	0.8160	0.0020	0.8001	0.0411	0.0194	73	0.39	032.42	0
032	Potassium (%)	0123	0.8200	0.0000	0.8001	0.0411	0.0194	73	0.48	032.41	0
032	Potassium (%)	0171	0.8200	0.0000	0.8001	0.0411	0.0194	73	0.48	032.41	0
032	Potassium (%)	0692	0.8200	0.0400	0.8001	0.0411	0.0194	73	0.48	032.42	0
032	Potassium (%)	0939	0.8200	0.0400	0.8001	0.0411	0.0194	73	0.48	032.53	0
032	Potassium (%)	0948	0.8235	0.0310	0.8001	0.0411	0.0194	73	0.57	032.31	0
032	Potassium (%)	0968	0.8260	0.0020	0.8001	0.0411	0.0194	73	0.63	032.43	0
032	Potassium (%)	0598	0.8295	0.0321	0.8001	0.0411	0.0194	73	0.71	032.41	0
032	Potassium (%)	0520	0.8300	0.0200	0.8001	0.0411	0.0194	73	0.73	032.41	0
032	Potassium (%)	0148	0.8325	0.0090	0.8001	0.0411	0.0194	73	0.79	032.41	0
032	Potassium (%)	0407	0.8347	0.0482	0.8001	0.0411	0.0194	73	0.84	032.41	0
032	Potassium (%)	0089	0.8400	0.0000	0.8001	0.0411	0.0194	73	0.97	032.43	0
032	Potassium (%)	0407	0.8422	0.0358	0.8001	0.0411	0.0194	73	1.02	032.43	0
032	Potassium (%)	0045	0.8440	0.0720	0.8001	0.0411	0.0194	73	1.07	032.42	0
032	Potassium (%)	0226	0.8450	0.0300	0.8001	0.0411	0.0194	73	1.09	032.41	0
032	Potassium (%)	0297	0.8450	0.0500	0.8001	0.0411	0.0194	73	1.09	032.43	0
032	Potassium (%)	0572	0.8490	0.0300	0.8001	0.0411	0.0194	73	1.19	032.53	0
032	Potassium (%)	0098	0.8500	0.0200	0.8001	0.0411	0.0194	73	1.21	032.43	0
032	Potassium (%)	0202	0.8560	0.0000	0.8001	0.0411	0.0194	73	1.36	032.43	0
032	Potassium (%)	0190	0.8850	0.0100	0.8001	0.0411	0.0194	73	2.06	032.42	0
032	Potassium (%)	0098	0.8900	0.0000	0.8001	0.0411	0.0194	73	2.19	032.41	0
032	Potassium (%)	0964	0.8910	0.0331	0.8001	0.0411	0.0194	73	2.21	032.43	0
032	Potassium (%)	0510	0.9000	0.0000	0.8001	0.0411	0.0194	73	2.43	032.43	0
032	Potassium (%)	0263	0.9225	0.0030	0.8001	0.0411	0.0194	73	2.98	032.42	0
032	Potassium (%)	0726	0.9250	0.0100	0.8001	0.0411	0.0194	73	3.04	032.42	0
033	Salt as chloride (%)	2246	0.5600	0.0000	1.109	0.0872	0.0226	67	-6.30	033.99	0
033	Salt as chloride (%)	0904	0.6676	0.0121	1.109	0.0872	0.0226	67	-5.06	033.05	0
033	Salt as chloride (%)	0536	0.7150	0.0100	1.109	0.0872	0.0226	67	-4.52	033.99	0
033	Salt as chloride (%)	0726	0.8200	0.1200	1.109	0.0872	0.0226	67	-3.31	033.03	0
033	Salt as chloride (%)	2234	0.8900	0.0600	1.109	0.0872	0.0226	67	-2.51	033.99	0
033	Salt as chloride (%)	2246	0.9150	0.0900	1.109	0.0872	0.0226	67	-2.23	033.00	0
033	Salt as chloride (%)	0358	0.9300	0.0400	1.109	0.0872	0.0226	67	-2.05	033.99	0
033	Salt as chloride (%)	0505	0.9400	0.0400	1.109	0.0872	0.0226	67	-1.94	033.03	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
033	Salt as chloride (%)	0353	0.9500	0.0200	1.109	0.0872	0.0226	67	-1.82	033.00	0
033	Salt as chloride (%)	2109	0.9550	0.0440	1.109	0.0872	0.0226	67	-1.77	033.01	0
033	Salt as chloride (%)	0265	0.9750	0.0300	1.109	0.0872	0.0226	67	-1.54	033.03	0
033	Salt as chloride (%)	0893	1.000	0.0000	1.109	0.0872	0.0226	67	-1.25	033.00	0
033	Salt as chloride (%)	2295	1.015	0.0100	1.109	0.0872	0.0226	67	-1.08	033.00	0
033	Salt as chloride (%)	0190	1.040	0.0000	1.109	0.0872	0.0226	67	-0.79	033.03	0
033	Salt as chloride (%)	2181	1.050	0.0800	1.109	0.0872	0.0226	67	-0.68	033.99	0
033	Salt as chloride (%)	0171	1.060	0.0000	1.109	0.0872	0.0226	67	-0.56	033.00	0
033	Salt as chloride (%)	0539	1.060	0.0200	1.109	0.0872	0.0226	67	-0.56	033.00	0
033	Salt as chloride (%)	0897	1.060	0.0200	1.109	0.0872	0.0226	67	-0.56	033.00	0
033	Salt as chloride (%)	2009	1.069	0.0001	1.109	0.0872	0.0226	67	-0.46	033.00	0
033	Salt as chloride (%)	2073	1.070	0.0000	1.109	0.0872	0.0226	67	-0.45	033.99	0
033	Salt as chloride (%)	2076	1.080	0.1000	1.109	0.0872	0.0226	67	-0.33	033.00	0
033	Salt as chloride (%)	0619	1.085	0.0100	1.109	0.0872	0.0226	67	-0.28	033.01	0
033	Salt as chloride (%)	2022	1.090	0.0000	1.109	0.0872	0.0226	67	-0.22	033.00	0
033	Salt as chloride (%)	0910	1.095	0.0500	1.109	0.0872	0.0226	67	-0.16	033.99	0
033	Salt as chloride (%)	2146	1.110	0.0000	1.109	0.0872	0.0226	67	0.01	033.00	0
033	Salt as chloride (%)	0083	1.110	0.0000	1.109	0.0872	0.0226	67	0.01	033.01	0
033	Salt as chloride (%)	0682	1.110	0.0000	1.109	0.0872	0.0226	67	0.01	033.99	0
033	Salt as chloride (%)	2062	1.115	0.0117	1.109	0.0872	0.0226	67	0.07	033.00	0
033	Salt as chloride (%)	0123	1.119	0.0040	1.109	0.0872	0.0226	67	0.11	033.00	0
033	Salt as chloride (%)	2196	1.120	0.0000	1.109	0.0872	0.0226	67	0.13	033.00	0
033	Salt as chloride (%)	0425	1.120	0.0000	1.109	0.0872	0.0226	67	0.13	033.01	0
033	Salt as chloride (%)	0881	1.123	0.0170	1.109	0.0872	0.0226	67	0.15	033.00	0
033	Salt as chloride (%)	0226	1.130	0.0000	1.109	0.0872	0.0226	67	0.24	033.01	0
033	Salt as chloride (%)	0510	1.130	0.0200	1.109	0.0872	0.0226	67	0.24	033.01	0
033	Salt as chloride (%)	0354	1.135	0.0100	1.109	0.0872	0.0226	67	0.30	033.01	0
033	Salt as chloride (%)	0885	1.135	0.0022	1.109	0.0872	0.0226	67	0.30	033.00	0
033	Salt as chloride (%)	0164	1.140	0.0000	1.109	0.0872	0.0226	67	0.35	033.01	0
033	Salt as chloride (%)	0939	1.140	0.0200	1.109	0.0872	0.0226	67	0.35	033.01	0
033	Salt as chloride (%)	0045	1.145	0.0300	1.109	0.0872	0.0226	67	0.41	033.00	0
033	Salt as chloride (%)	0878	1.145	0.0100	1.109	0.0872	0.0226	67	0.41	033.00	0
033	Salt as chloride (%)	0683	1.145	0.0100	1.109	0.0872	0.0226	67	0.41	033.01	0
033	Salt as chloride (%)	0689	1.145	0.0100	1.109	0.0872	0.0226	67	0.41	033.05	0
033	Salt as chloride (%)	0001	1.150	0.0200	1.109	0.0872	0.0226	67	0.47	033.01	0
033	Salt as chloride (%)	0003	1.150	0.1000	1.109	0.0872	0.0226	67	0.47	033.01	0
033	Salt as chloride (%)	0242	1.150	0.0200	1.109	0.0872	0.0226	67	0.47	033.01	0
033	Salt as chloride (%)	0948	1.150	0.0200	1.109	0.0872	0.0226	67	0.47	033.01	0
033	Salt as chloride (%)	2146	1.150	0.0000	1.109	0.0872	0.0226	67	0.47	033.01	0
033	Salt as chloride (%)	0868	1.150	0.0200	1.109	0.0872	0.0226	67	0.47	033.05	0
033	Salt as chloride (%)	0407	1.152	0.0170	1.109	0.0872	0.0226	67	0.49	033.01	0
033	Salt as chloride (%)	0964	1.153	0.0550	1.109	0.0872	0.0226	67	0.50	033.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
033	Salt as chloride (%)	0004	1.155	0.0100	1.109	0.0872	0.0226	67	0.53	033.01	0
033	Salt as chloride (%)	0098	1.155	0.0100	1.109	0.0872	0.0226	67	0.53	033.01	0
033	Salt as chloride (%)	0100	1.155	0.0300	1.109	0.0872	0.0226	67	0.53	033.01	0
033	Salt as chloride (%)	0229	1.155	0.0100	1.109	0.0872	0.0226	67	0.53	033.01	0
033	Salt as chloride (%)	0298	1.160	0.0200	1.109	0.0872	0.0226	67	0.58	033.00	0
033	Salt as chloride (%)	0175	1.160	0.0000	1.109	0.0872	0.0226	67	0.58	033.01	0
033	Salt as chloride (%)	0169	1.165	0.0100	1.109	0.0872	0.0226	67	0.64	033.00	0
033	Salt as chloride (%)	0590	1.165	0.0100	1.109	0.0872	0.0226	67	0.64	033.01	0
033	Salt as chloride (%)	0309	1.167	0.0209	1.109	0.0872	0.0226	67	0.66	033.00	0
033	Salt as chloride (%)	0019	1.180	0.0000	1.109	0.0872	0.0226	67	0.81	033.01	0
033	Salt as chloride (%)	0693	1.223	0.0710	1.109	0.0872	0.0226	67	1.30	033.00	0
033	Salt as chloride (%)	0278	1.225	0.0100	1.109	0.0872	0.0226	67	1.33	033.01	0
033	Salt as chloride (%)	0186	1.228	0.0160	1.109	0.0872	0.0226	67	1.36	033.01	0
033	Salt as chloride (%)	0675	1.285	0.1300	1.109	0.0872	0.0226	67	2.02	033.00	0
033	Salt as chloride (%)	2268	1.290	0.0000	1.109	0.0872	0.0226	67	2.07	033.01	0
033	Salt as chloride (%)	0144	1.310	0.0000	1.109	0.0872	0.0226	67	2.30	033.03	0
033	Salt as chloride (%)	0148	1.905	0.0100	1.109	0.0872	0.0226	67	9.13	033.01	0
033	Salt as chloride (%)	2129	1.059	0.2327	1.109	0.0872	0.0226	67	-0.57	033.99	1
033	Salt as chloride (%)	2043	1.507	0.2050	1.109	0.0872	0.0226	67	4.56	033.05	1
034	Selenium (ppm)	0066	0.3200	0.0200	1.172	0.2072	0.0537	24	-4.11	034.43	0
034	Selenium (ppm)	0692	0.7500	0.0000	1.172	0.2072	0.0537	24	-2.04	034.42	0
034	Selenium (ppm)	0619	0.9460	0.0400	1.172	0.2072	0.0537	24	-1.09	034.41	0
034	Selenium (ppm)	0045	0.9700	0.0020	1.172	0.2072	0.0537	24	-0.97	034.04	0
034	Selenium (ppm)	0407	1.015	0.0900	1.172	0.2072	0.0537	24	-0.76	034.41	0
034	Selenium (ppm)	0164	1.030	0.1010	1.172	0.2072	0.0537	24	-0.69	034.53	0
034	Selenium (ppm)	0047	1.045	0.0300	1.172	0.2072	0.0537	24	-0.61	034.52	0
034	Selenium (ppm)	0148	1.059	0.1430	1.172	0.2072	0.0537	24	-0.55	034.41	0
034	Selenium (ppm)	0914	1.086	0.0750	1.172	0.2072	0.0537	24	-0.42	034.53	0
034	Selenium (ppm)	0171	1.116	0.0030	1.172	0.2072	0.0537	24	-0.27	034.04	0
034	Selenium (ppm)	0553	1.150	0.0200	1.172	0.2072	0.0537	24	-0.11	034.53	0
034	Selenium (ppm)	0910	1.155	0.0300	1.172	0.2072	0.0537	24	-0.08	034.52	0
034	Selenium (ppm)	0563	1.167	0.0411	1.172	0.2072	0.0537	24	-0.02	034.04	0
034	Selenium (ppm)	0682	1.200	0.0000	1.172	0.2072	0.0537	24	0.14	034.43	0
034	Selenium (ppm)	0186	1.228	0.0220	1.172	0.2072	0.0537	24	0.27	034.52	0
034	Selenium (ppm)	0169	1.255	0.0100	1.172	0.2072	0.0537	24	0.40	034.04	0
034	Selenium (ppm)	0560	1.268	0.0960	1.172	0.2072	0.0537	24	0.46	034.53	0
034	Selenium (ppm)	0964	1.275	0.0200	1.172	0.2072	0.0537	24	0.50	034.43	0
034	Selenium (ppm)	0939	1.310	0.0400	1.172	0.2072	0.0537	24	0.67	034.53	0
034	Selenium (ppm)	0918	1.339	0.0688	1.172	0.2072	0.0537	24	0.81	034.53	0
034	Selenium (ppm)	2033	1.345	0.0900	1.172	0.2072	0.0537	24	0.84	034.53	0
034	Selenium (ppm)	0968	1.605	0.0070	1.172	0.2072	0.0537	24	2.09	034.43	0
034	Selenium (ppm)	0407	1.715	0.3300	1.172	0.2072	0.0537	24	2.62	034.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
034	Selenium (ppm)	0190	2.605	0.0100	1.172	0.2072	0.0537	24	6.92	034.99	0
034	Selenium (ppm)	2234	1.175	0.3900	1.172	0.2072	0.0537	24	0.01	034.43	1
034	Selenium (ppm)	2302	361.0	0.0000	1.172	0.2072	0.0537	24	1736.79	034.99	2
035	Sodium (%)	0358	0.3650	0.0100	0.4110	0.0231	0.0094	77	-1.99	035.41	0
035	Sodium (%)	0278	0.3650	0.0100	0.4110	0.0231	0.0094	77	-1.99	035.42	0
035	Sodium (%)	2012	0.3700	0.0000	0.4110	0.0231	0.0094	77	-1.77	035.41	0
035	Sodium (%)	2129	0.3728	0.0005	0.4110	0.0231	0.0094	77	-1.66	035.42	0
035	Sodium (%)	0683	0.3750	0.0100	0.4110	0.0231	0.0094	77	-1.56	035.31	0
035	Sodium (%)	0144	0.3750	0.0100	0.4110	0.0231	0.0094	77	-1.56	035.41	0
035	Sodium (%)	0675	0.3750	0.0100	0.4110	0.0231	0.0094	77	-1.56	035.43	0
035	Sodium (%)	2259	0.3825	0.0270	0.4110	0.0231	0.0094	77	-1.23	035.41	0
035	Sodium (%)	0939	0.3850	0.0100	0.4110	0.0231	0.0094	77	-1.13	035.31	0
035	Sodium (%)	0563	0.3896	0.0052	0.4110	0.0231	0.0094	77	-0.93	035.31	0
035	Sodium (%)	2196	0.3900	0.0000	0.4110	0.0231	0.0094	77	-0.91	035.31	0
035	Sodium (%)	0019	0.3900	0.0200	0.4110	0.0231	0.0094	77	-0.91	035.41	0
035	Sodium (%)	0298	0.3900	0.0200	0.4110	0.0231	0.0094	77	-0.91	035.41	0
035	Sodium (%)	0265	0.3900	0.0200	0.4110	0.0231	0.0094	77	-0.91	035.42	0
035	Sodium (%)	0870	0.3922	0.0229	0.4110	0.0231	0.0094	77	-0.82	035.42	0
035	Sodium (%)	0689	0.3950	0.0300	0.4110	0.0231	0.0094	77	-0.69	035.31	0
035	Sodium (%)	0229	0.3950	0.0100	0.4110	0.0231	0.0094	77	-0.69	035.41	0
035	Sodium (%)	0035	0.3950	0.0100	0.4110	0.0231	0.0094	77	-0.69	035.42	0
035	Sodium (%)	2192	0.3950	0.0100	0.4110	0.0231	0.0094	77	-0.69	035.43	0
035	Sodium (%)	0242	0.3950	0.0100	0.4110	0.0231	0.0094	77	-0.69	035.99	0
035	Sodium (%)	0560	0.3956	0.0114	0.4110	0.0231	0.0094	77	-0.67	035.52	0
035	Sodium (%)	0037	0.3985	0.0110	0.4110	0.0231	0.0094	77	-0.54	035.42	0
035	Sodium (%)	0868	0.3995	0.0030	0.4110	0.0231	0.0094	77	-0.50	035.01	0
035	Sodium (%)	0148	0.3995	0.0110	0.4110	0.0231	0.0094	77	-0.50	035.41	0
035	Sodium (%)	0536	0.4000	0.0000	0.4110	0.0231	0.0094	77	-0.48	035.31	0
035	Sodium (%)	0297	0.4000	0.0000	0.4110	0.0231	0.0094	77	-0.48	035.43	0
035	Sodium (%)	0142	0.4012	0.0024	0.4110	0.0231	0.0094	77	-0.42	035.31	0
035	Sodium (%)	0948	0.4020	0.0080	0.4110	0.0231	0.0094	77	-0.39	035.31	0
035	Sodium (%)	0345	0.4025	0.0090	0.4110	0.0231	0.0094	77	-0.37	035.43	0
035	Sodium (%)	0693	0.4035	0.0010	0.4110	0.0231	0.0094	77	-0.32	035.42	0
035	Sodium (%)	0171	0.4050	0.0100	0.4110	0.0231	0.0094	77	-0.26	035.41	0
035	Sodium (%)	0910	0.4050	0.0100	0.4110	0.0231	0.0094	77	-0.26	035.41	0
035	Sodium (%)	0035	0.4050	0.0100	0.4110	0.0231	0.0094	77	-0.26	035.43	0
035	Sodium (%)	0918	0.4055	0.0010	0.4110	0.0231	0.0094	77	-0.24	035.43	0
035	Sodium (%)	0529	0.4100	0.0040	0.4110	0.0231	0.0094	77	-0.04	035.31	0
035	Sodium (%)	0656	0.4100	0.0000	0.4110	0.0231	0.0094	77	-0.04	035.31	0
035	Sodium (%)	0354	0.4100	0.0000	0.4110	0.0231	0.0094	77	-0.04	035.41	0
035	Sodium (%)	0726	0.4100	0.0000	0.4110	0.0231	0.0094	77	-0.04	035.42	0
035	Sodium (%)	0425	0.4100	0.0000	0.4110	0.0231	0.0094	77	-0.04	035.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
035	Sodium (%)	0682	0.4100	0.0000	0.4110	0.0231	0.0094	77	-0.04	035.43	0
035	Sodium (%)	0553	0.4100	0.0220	0.4110	0.0231	0.0094	77	-0.04	035.53	0
035	Sodium (%)	0083	0.4108	0.0007	0.4110	0.0231	0.0094	77	-0.01	035.43	0
035	Sodium (%)	0186	0.4128	0.0002	0.4110	0.0231	0.0094	77	0.08	035.52	0
035	Sodium (%)	0154	0.4136	0.0043	0.4110	0.0231	0.0094	77	0.11	035.52	0
035	Sodium (%)	0870	0.4139	0.0214	0.4110	0.0231	0.0094	77	0.13	035.43	0
035	Sodium (%)	0042	0.4140	0.0020	0.4110	0.0231	0.0094	77	0.13	035.42	0
035	Sodium (%)	0353	0.4140	0.0124	0.4110	0.0231	0.0094	77	0.13	035.43	0
035	Sodium (%)	0868	0.4150	0.0100	0.4110	0.0231	0.0094	77	0.17	035.31	0
035	Sodium (%)	0878	0.4150	0.0100	0.4110	0.0231	0.0094	77	0.17	035.41	0
035	Sodium (%)	0100	0.4150	0.0100	0.4110	0.0231	0.0094	77	0.17	035.99	0
035	Sodium (%)	0169	0.4150	0.0300	0.4110	0.0231	0.0094	77	0.17	035.32	0
035	Sodium (%)	0572	0.4215	0.0190	0.4110	0.0231	0.0094	77	0.46	035.53	0
035	Sodium (%)	0045	0.4220	0.0420	0.4110	0.0231	0.0094	77	0.48	035.42	0
035	Sodium (%)	0520	0.4232	0.0008	0.4110	0.0231	0.0094	77	0.53	035.41	0
035	Sodium (%)	0164	0.4240	0.0040	0.4110	0.0231	0.0094	77	0.56	035.41	0
035	Sodium (%)	0190	0.4250	0.0100	0.4110	0.0231	0.0094	77	0.61	035.42	0
035	Sodium (%)	0098	0.4250	0.0100	0.4110	0.0231	0.0094	77	0.61	035.43	0
035	Sodium (%)	0619	0.4255	0.0030	0.4110	0.0231	0.0094	77	0.63	035.41	0
035	Sodium (%)	0263	0.4260	0.0020	0.4110	0.0231	0.0094	77	0.65	035.42	0
035	Sodium (%)	0407	0.4275	0.0008	0.4110	0.0231	0.0094	77	0.72	035.41	0
035	Sodium (%)	0881	0.4285	0.0150	0.4110	0.0231	0.0094	77	0.76	035.43	0
035	Sodium (%)	0108	0.4300	0.0000	0.4110	0.0231	0.0094	77	0.82	035.05	0
035	Sodium (%)	0692	0.4300	0.0000	0.4110	0.0231	0.0094	77	0.82	035.42	0
035	Sodium (%)	0407	0.4348	0.0155	0.4110	0.0231	0.0094	77	1.03	035.43	0
035	Sodium (%)	0123	0.4350	0.0100	0.4110	0.0231	0.0094	77	1.04	035.41	0
035	Sodium (%)	0590	0.4350	0.0300	0.4110	0.0231	0.0094	77	1.04	035.99	0
035	Sodium (%)	0202	0.4390	0.0000	0.4110	0.0231	0.0094	77	1.21	035.43	0
035	Sodium (%)	0089	0.4400	0.0000	0.4110	0.0231	0.0094	77	1.26	035.43	0
035	Sodium (%)	0598	0.4425	0.0044	0.4110	0.0231	0.0094	77	1.36	035.41	0
035	Sodium (%)	0186	0.4433	0.0037	0.4110	0.0231	0.0094	77	1.40	035.42	0
035	Sodium (%)	0098	0.4450	0.0300	0.4110	0.0231	0.0094	77	1.47	035.41	0
035	Sodium (%)	0294	0.4450	0.0100	0.4110	0.0231	0.0094	77	1.47	035.42	0
035	Sodium (%)	0510	0.4450	0.0100	0.4110	0.0231	0.0094	77	1.47	035.43	0
035	Sodium (%)	0964	0.4520	0.0186	0.4110	0.0231	0.0094	77	1.78	035.43	0
035	Sodium (%)	0175	0.4600	0.0000	0.4110	0.0231	0.0094	77	2.12	035.31	0
035	Sodium (%)	0226	0.4800	0.0000	0.4110	0.0231	0.0094	77	2.99	035.41	0
035	Sodium (%)	0038	0.4920	0.0280	0.4110	0.0231	0.0094	77	3.51	035.43	0
035	Sodium (%)	0939	0.4150	0.0500	0.4110	0.0231	0.0094	77	0.17	035.53	1
035	Sodium (%)	2302	0.2100	0.0000	0.4110	0.0231	0.0094	77	-8.70	035.99	2
036	Sulfur (%)	0870	0.2184	0.0103	0.2856	0.0182	0.0105	42	-3.70	036.42	0
036	Sulfur (%)	2033	0.2300	0.0000	0.2856	0.0182	0.0105	42	-3.06	036.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
036	Sulfur (%)	0226	0.2600	0.0000	0.2856	0.0182	0.0105	42	-1.41	036.04	0
036	Sulfur (%)	0692	0.2600	0.0200	0.2856	0.0182	0.0105	42	-1.41	036.42	0
036	Sulfur (%)	0870	0.2631	0.0115	0.2856	0.0182	0.0105	42	-1.24	036.43	0
036	Sulfur (%)	0693	0.2640	0.0060	0.2856	0.0182	0.0105	42	-1.19	036.42	0
036	Sulfur (%)	0357	0.2672	0.0023	0.2856	0.0182	0.0105	42	-1.02	036.42	0
036	Sulfur (%)	0353	0.2700	0.0200	0.2856	0.0182	0.0105	42	-0.86	036.43	0
036	Sulfur (%)	0242	0.2700	0.0200	0.2856	0.0182	0.0105	42	-0.86	036.99	0
036	Sulfur (%)	0910	0.2723	0.0315	0.2856	0.0182	0.0105	42	-0.74	036.52	0
036	Sulfur (%)	0407	0.2736	0.0020	0.2856	0.0182	0.0105	42	-0.66	036.43	0
036	Sulfur (%)	0045	0.2750	0.0140	0.2856	0.0182	0.0105	42	-0.58	036.42	0
036	Sulfur (%)	0278	0.2750	0.0100	0.2856	0.0182	0.0105	42	-0.58	036.42	0
036	Sulfur (%)	2129	0.2757	0.0052	0.2856	0.0182	0.0105	42	-0.54	036.42	0
036	Sulfur (%)	0148	0.2800	0.0000	0.2856	0.0182	0.0105	42	-0.31	036.04	0
036	Sulfur (%)	0035	0.2800	0.0200	0.2856	0.0182	0.0105	42	-0.31	036.43	0
036	Sulfur (%)	0297	0.2800	0.0200	0.2856	0.0182	0.0105	42	-0.31	036.43	0
036	Sulfur (%)	0553	0.2845	0.0370	0.2856	0.0182	0.0105	42	-0.06	036.53	0
036	Sulfur (%)	0265	0.2850	0.0100	0.2856	0.0182	0.0105	42	-0.03	036.42	0
036	Sulfur (%)	0294	0.2850	0.0100	0.2856	0.0182	0.0105	42	-0.03	036.42	0
036	Sulfur (%)	0083	0.2850	0.0100	0.2856	0.0182	0.0105	42	-0.03	036.43	0
036	Sulfur (%)	0169	0.2850	0.0100	0.2856	0.0182	0.0105	42	-0.03	036.43	0
036	Sulfur (%)	0186	0.2890	0.0074	0.2856	0.0182	0.0105	42	0.19	036.42	0
036	Sulfur (%)	0345	0.2895	0.0030	0.2856	0.0182	0.0105	42	0.22	036.43	0
036	Sulfur (%)	0164	0.2900	0.0000	0.2856	0.0182	0.0105	42	0.24	036.42	0
036	Sulfur (%)	0229	0.2900	0.0000	0.2856	0.0182	0.0105	42	0.24	036.42	0
036	Sulfur (%)	0726	0.2900	0.0000	0.2856	0.0182	0.0105	42	0.24	036.42	0
036	Sulfur (%)	0186	0.2925	0.0005	0.2856	0.0182	0.0105	42	0.38	036.52	0
036	Sulfur (%)	0918	0.2935	0.0030	0.2856	0.0182	0.0105	42	0.44	036.43	0
036	Sulfur (%)	0190	0.2950	0.0100	0.2856	0.0182	0.0105	42	0.52	036.42	0
036	Sulfur (%)	0708	0.2970	0.0020	0.2856	0.0182	0.0105	42	0.63	036.42	0
036	Sulfur (%)	0038	0.2975	0.0150	0.2856	0.0182	0.0105	42	0.66	036.43	0
036	Sulfur (%)	0037	0.2990	0.0000	0.2856	0.0182	0.0105	42	0.74	036.42	0
036	Sulfur (%)	0407	0.2991	0.0183	0.2856	0.0182	0.0105	42	0.74	036.42	0
036	Sulfur (%)	0035	0.3000	0.0200	0.2856	0.0182	0.0105	42	0.79	036.42	0
036	Sulfur (%)	0202	0.3050	0.0080	0.2856	0.0182	0.0105	42	1.07	036.43	0
036	Sulfur (%)	0042	0.3055	0.0110	0.2856	0.0182	0.0105	42	1.10	036.42	0
036	Sulfur (%)	0560	0.3069	0.0352	0.2856	0.0182	0.0105	42	1.17	036.52	0
036	Sulfur (%)	0171	0.3100	0.0000	0.2856	0.0182	0.0105	42	1.34	036.42	0
036	Sulfur (%)	0510	0.3300	0.0000	0.2856	0.0182	0.0105	42	2.45	036.43	0
036	Sulfur (%)	0098	0.3326	0.0060	0.2856	0.0182	0.0105	42	2.59	036.43	0
036	Sulfur (%)	0229	0.4550	0.0300	0.2856	0.0182	0.0105	42	9.33	036.04	0
037	Zinc (ppm)	0047	275.6	7.500	388.6	26.65	12.30	86	-4.24	037.52	0
037	Zinc (ppm)	2129	337.6	27.90	388.6	26.65	12.30	86	-1.91	037.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
037	Zinc (ppm)	0294	345.6	5.270	388.6	26.65	12.30	86	-1.61	037.42	0
037	Zinc (ppm)	0354	345.7	2.230	388.6	26.65	12.30	86	-1.61	037.41	0
037	Zinc (ppm)	0619	346.5	2.367	388.6	26.65	12.30	86	-1.58	037.41	0
037	Zinc (ppm)	0675	350.2	0.1200	388.6	26.65	12.30	86	-1.44	037.43	0
037	Zinc (ppm)	0921	353.0	6.000	388.6	26.65	12.30	86	-1.33	037.31	0
037	Zinc (ppm)	0910	355.0	28.00	388.6	26.65	12.30	86	-1.26	037.41	0
037	Zinc (ppm)	0278	355.1	0.7000	388.6	26.65	12.30	86	-1.26	037.42	0
037	Zinc (ppm)	0066	358.7	8.770	388.6	26.65	12.30	86	-1.12	037.44	0
037	Zinc (ppm)	0870	359.4	4.500	388.6	26.65	12.30	86	-1.10	037.42	0
037	Zinc (ppm)	0045	359.5	73.00	388.6	26.65	12.30	86	-1.09	037.42	0
037	Zinc (ppm)	0357	360.7	10.39	388.6	26.65	12.30	86	-1.05	037.42	0
037	Zinc (ppm)	0186	361.0	22.00	388.6	26.65	12.30	86	-1.03	037.42	0
037	Zinc (ppm)	2302	361.0	0.0000	388.6	26.65	12.30	86	-1.03	037.99	0
037	Zinc (ppm)	0175	362.0	8.000	388.6	26.65	12.30	86	-1.00	037.31	0
037	Zinc (ppm)	2234	364.0	10.57	388.6	26.65	12.30	86	-0.92	037.43	0
037	Zinc (ppm)	0529	365.7	8.300	388.6	26.65	12.30	86	-0.86	037.31	0
037	Zinc (ppm)	0038	369.5	61.00	388.6	26.65	12.30	86	-0.72	037.32	0
037	Zinc (ppm)	0693	371.5	10.55	388.6	26.65	12.30	86	-0.64	037.42	0
037	Zinc (ppm)	0042	372.0	10.00	388.6	26.65	12.30	86	-0.62	037.42	0
037	Zinc (ppm)	2152	373.4	0.6200	388.6	26.65	12.30	86	-0.57	037.44	0
037	Zinc (ppm)	0692	374.5	41.00	388.6	26.65	12.30	86	-0.53	037.42	0
037	Zinc (ppm)	0918	374.8	4.610	388.6	26.65	12.30	86	-0.52	037.43	0
037	Zinc (ppm)	2012	376.6	0.7000	388.6	26.65	12.30	86	-0.45	037.41	0
037	Zinc (ppm)	0948	377.4	3.170	388.6	26.65	12.30	86	-0.42	037.33	0
037	Zinc (ppm)	2192	377.5	29.65	388.6	26.65	12.30	86	-0.42	037.43	0
037	Zinc (ppm)	0345	377.9	2.700	388.6	26.65	12.30	86	-0.40	037.43	0
037	Zinc (ppm)	0263	378.0	1.797	388.6	26.65	12.30	86	-0.40	037.42	0
037	Zinc (ppm)	0001	378.4	9.300	388.6	26.65	12.30	86	-0.38	037.31	0
037	Zinc (ppm)	0169	378.5	17.00	388.6	26.65	12.30	86	-0.38	037.43	0
037	Zinc (ppm)	0870	378.6	9.500	388.6	26.65	12.30	86	-0.38	037.43	0
037	Zinc (ppm)	0563	379.1	3.918	388.6	26.65	12.30	86	-0.36	037.31	0
037	Zinc (ppm)	2188	379.3	0.7200	388.6	26.65	12.30	86	-0.35	037.31	0
037	Zinc (ppm)	0353	380.5	7.000	388.6	26.65	12.30	86	-0.30	037.43	0
037	Zinc (ppm)	0536	380.8	0.2000	388.6	26.65	12.30	86	-0.29	037.31	0
037	Zinc (ppm)	0572	381.0	10.00	388.6	26.65	12.30	86	-0.28	037.53	0
037	Zinc (ppm)	0358	381.4	32.03	388.6	26.65	12.30	86	-0.27	037.41	0
037	Zinc (ppm)	0148	382.5	3.000	388.6	26.65	12.30	86	-0.23	037.41	0
037	Zinc (ppm)	0689	383.2	16.20	388.6	26.65	12.30	86	-0.20	037.31	0
037	Zinc (ppm)	0242	383.5	25.00	388.6	26.65	12.30	86	-0.19	037.99	0
037	Zinc (ppm)	2033	384.3	1.000	388.6	26.65	12.30	86	-0.16	037.43	0
037	Zinc (ppm)	0074	384.5	21.00	388.6	26.65	12.30	86	-0.15	037.41	0
037	Zinc (ppm)	0004	384.8	61.74	388.6	26.65	12.30	86	-0.14	037.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
037	Zinc (ppm)	2196	386.1	0.0000	388.6	26.65	12.30	86	-0.09	037.31	0
037	Zinc (ppm)	0560	386.5	41.40	388.6	26.65	12.30	86	-0.08	037.52	0
037	Zinc (ppm)	0939	387.5	7.000	388.6	26.65	12.30	86	-0.04	037.53	0
037	Zinc (ppm)	0682	388.3	0.0000	388.6	26.65	12.30	86	-0.01	037.43	0
037	Zinc (ppm)	0164	389.0	4.000	388.6	26.65	12.30	86	0.02	037.44	0
037	Zinc (ppm)	0171	393.0	6.000	388.6	26.65	12.30	86	0.17	037.41	0
037	Zinc (ppm)	0868	395.5	1.000	388.6	26.65	12.30	86	0.26	037.31	0
037	Zinc (ppm)	0878	396.0	2.000	388.6	26.65	12.30	86	0.28	037.41	0
037	Zinc (ppm)	0968	396.0	2.000	388.6	26.65	12.30	86	0.28	037.43	0
037	Zinc (ppm)	0407	396.3	4.715	388.6	26.65	12.30	86	0.29	037.41	0
037	Zinc (ppm)	0083	400.0	20.00	388.6	26.65	12.30	86	0.43	037.43	0
037	Zinc (ppm)	0510	400.0	0.0000	388.6	26.65	12.30	86	0.43	037.43	0
037	Zinc (ppm)	0425	400.2	45.40	388.6	26.65	12.30	86	0.44	037.43	0
037	Zinc (ppm)	0939	401.9	2.540	388.6	26.65	12.30	86	0.50	037.31	0
037	Zinc (ppm)	0003	402.0	20.00	388.6	26.65	12.30	86	0.50	037.41	0
037	Zinc (ppm)	0098	402.7	16.00	388.6	26.65	12.30	86	0.53	037.41	0
037	Zinc (ppm)	0019	404.0	0.1000	388.6	26.65	12.30	86	0.58	037.41	0
037	Zinc (ppm)	0186	404.0	26.00	388.6	26.65	12.30	86	0.58	037.52	0
037	Zinc (ppm)	0010	405.0	10.00	388.6	26.65	12.30	86	0.62	037.33	0
037	Zinc (ppm)	0505	405.2	13.00	388.6	26.65	12.30	86	0.62	037.43	0
037	Zinc (ppm)	0656	405.8	0.0000	388.6	26.65	12.30	86	0.65	037.31	0
037	Zinc (ppm)	2144	408.7	6.960	388.6	26.65	12.30	86	0.75	037.99	0
037	Zinc (ppm)	0098	410.4	16.70	388.6	26.65	12.30	86	0.82	037.43	0
037	Zinc (ppm)	0520	411.0	2.000	388.6	26.65	12.30	86	0.84	037.41	0
037	Zinc (ppm)	0848	411.9	1.720	388.6	26.65	12.30	86	0.88	037.41	0
037	Zinc (ppm)	0726	413.6	0.0000	388.6	26.65	12.30	86	0.94	037.42	0
037	Zinc (ppm)	0229	416.3	14.32	388.6	26.65	12.30	86	1.04	037.41	0
037	Zinc (ppm)	0265	416.5	39.00	388.6	26.65	12.30	86	1.05	037.42	0
037	Zinc (ppm)	0598	417.3	2.800	388.6	26.65	12.30	86	1.08	037.41	0
037	Zinc (ppm)	0202	417.7	11.20	388.6	26.65	12.30	86	1.09	037.43	0
037	Zinc (ppm)	0190	417.9	4.100	388.6	26.65	12.30	86	1.10	037.42	0
037	Zinc (ppm)	0035	420.5	1.000	388.6	26.65	12.30	86	1.20	037.42	0
037	Zinc (ppm)	0297	421.5	15.00	388.6	26.65	12.30	86	1.24	037.43	0
037	Zinc (ppm)	0035	422.0	16.00	388.6	26.65	12.30	86	1.25	037.43	0
037	Zinc (ppm)	0553	423.5	29.00	388.6	26.65	12.30	86	1.31	037.53	0
037	Zinc (ppm)	0590	423.5	5.000	388.6	26.65	12.30	86	1.31	037.99	0
037	Zinc (ppm)	0407	426.1	4.489	388.6	26.65	12.30	86	1.41	037.43	0
037	Zinc (ppm)	0038	435.0	6.000	388.6	26.65	12.30	86	1.74	037.43	0
037	Zinc (ppm)	0904	439.0	27.05	388.6	26.65	12.30	86	1.89	037.33	0
037	Zinc (ppm)	0100	443.5	7.000	388.6	26.65	12.30	86	2.06	037.99	0
037	Zinc (ppm)	0914	448.2	11.20	388.6	26.65	12.30	86	2.24	037.43	0
037	Zinc (ppm)	0037	464.7	6.800	388.6	26.65	12.30	86	2.86	037.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
037	Zinc (ppm)	0226	17,893	35,032	388.6	26.65	12.30	86	656.88	037.41	2
038	Molybdenum (ppm)	0353	0.7245	0.0279	1.087	0.1844	0.0576	19	-1.96	038.43	0
038	Molybdenum (ppm)	0171	0.8130	0.0080	1.087	0.1844	0.0576	19	-1.48	038.41	0
038	Molybdenum (ppm)	0345	0.8700	0.0200	1.087	0.1844	0.0576	19	-1.17	038.43	0
038	Molybdenum (ppm)	0297	0.9200	0.0600	1.087	0.1844	0.0576	19	-0.90	038.43	0
038	Molybdenum (ppm)	0278	0.9450	0.0900	1.087	0.1844	0.0576	19	-0.77	038.42	0
038	Molybdenum (ppm)	0910	1.040	0.1000	1.087	0.1844	0.0576	19	-0.25	038.52	0
038	Molybdenum (ppm)	0407	1.044	0.0429	1.087	0.1844	0.0576	19	-0.23	038.41	0
038	Molybdenum (ppm)	0407	1.070	0.0437	1.087	0.1844	0.0576	19	-0.09	038.43	0
038	Molybdenum (ppm)	0226	1.110	0.0400	1.087	0.1844	0.0576	19	0.13	038.41	0
038	Molybdenum (ppm)	0169	1.110	0.0400	1.087	0.1844	0.0576	19	0.13	038.43	0
038	Molybdenum (ppm)	0918	1.111	0.0410	1.087	0.1844	0.0576	19	0.13	038.53	0
038	Molybdenum (ppm)	0693	1.130	0.1610	1.087	0.1844	0.0576	19	0.23	038.42	0
038	Molybdenum (ppm)	2033	1.135	0.0500	1.087	0.1844	0.0576	19	0.26	038.43	0
038	Molybdenum (ppm)	0553	1.155	0.0300	1.087	0.1844	0.0576	19	0.37	038.53	0
038	Molybdenum (ppm)	0939	1.170	0.0400	1.087	0.1844	0.0576	19	0.45	038.53	0
038	Molybdenum (ppm)	0164	1.185	0.0100	1.087	0.1844	0.0576	19	0.53	038.53	0
038	Molybdenum (ppm)	0083	1.300	0.0000	1.087	0.1844	0.0576	19	1.16	038.43	0
038	Molybdenum (ppm)	0510	1.550	0.1000	1.087	0.1844	0.0576	19	2.51	038.43	0
038	Molybdenum (ppm)	0045	1.715	0.1900	1.087	0.1844	0.0576	19	3.41	038.42	0
038	Molybdenum (ppm)	0265	2.450	0.5000	1.087	0.1844	0.0576	19	7.39	038.42	1
040	Barium (ppm)	0918	8.020	0.0981				2		040.53	0
040	Barium (ppm)	0560	8.445	0.2100				2		040.52	0
041	Vanadium (ppm)	0553	1.310	0.0200				2		041.53	0
041	Vanadium (ppm)	2033	1.455	0.0500				2		041.43	0
042	Chloride (%)	0619	0.6560	0.0080				3		042.00	0
042	Chloride (%)	0948	0.6990	0.0120				3		042.00	0
042	Chloride (%)	0297	0.7150	0.0100				3		042.99	0
101	Choline Chloride (ppm)	0904	1,589	241.0				3		101.02	0
101	Choline Chloride (ppm)	0171	1,883	47.00				3		101.01	0
101	Choline Chloride (ppm)	0227	2,630	120.0				3		101.99	0
102	Niacin (ppm)	0160	34.45	7.500	52.52	21.93	4.170	4		102.99	0
102	Niacin (ppm)	0010	38.90	1.600	52.52	21.93	4.170	4		102.02	0
102	Niacin (ppm)	0904	53.71	4.580	52.52	21.93	4.170	4		102.99	0
102	Niacin (ppm)	0227	83.00	3.000	52.52	21.93	4.170	4		102.01	0
103	Pantothenic Acid (ppm)	0227	29.85	2.300				3		103.01	0
103	Pantothenic Acid (ppm)	0010	35.10	3.600				3		103.02	0
103	Pantothenic Acid (ppm)	0160	111.1	19.50				3		103.99	0
104	Riboflavin (ppm)	0160	5.180	0.1200	9.065	2.306	1.125	8	-1.68	104.99	0
104	Riboflavin (ppm)	0910	6.150	0.1000	9.065	2.306	1.125	8	-1.26	104.03	0
104	Riboflavin (ppm)	0027	9.392	1.058	9.065	2.306	1.125	8	0.14	104.03	0
104	Riboflavin (ppm)	2192	9.400	3.000	9.065	2.306	1.125	8	0.15	104.03	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
104	Riboflavin (ppm)	0939	10.30	0.2500	9.065	2.306	1.125	8	0.53	104.03	0
104	Riboflavin (ppm)	0904	10.38	2.270	9.065	2.306	1.125	8	0.57	104.99	0
104	Riboflavin (ppm)	0010	10.60	2.200	9.065	2.306	1.125	8	0.67	104.03	0
104	Riboflavin (ppm)	0227	10.70	0.0000	9.065	2.306	1.125	8	0.71	104.00	0
105	Thiamine (ppm)	0910	2.110	0.0800				3		105.00	0
105	Thiamine (ppm)	0227	2.825	0.0700				3		105.01	0
105	Thiamine (ppm)	0939	2.895	0.0500				3		105.00	0
106	Vitamin A (KU / kg)	0905	0.7470	0.0280	2.576	2.103	1.089	16	-0.87	106.02	0
106	Vitamin A (KU / kg)	0875	0.7600	0.0800	2.576	2.103	1.089	16	-0.86	106.02	0
106	Vitamin A (KU / kg)	0676	0.7950	0.0300	2.576	2.103	1.089	16	-0.85	106.02	0
106	Vitamin A (KU / kg)	0227	1.011	0.0990	2.576	2.103	1.089	16	-0.74	106.00	0
106	Vitamin A (KU / kg)	0148	1.310	0.0400	2.576	2.103	1.089	16	-0.60	106.02	0
106	Vitamin A (KU / kg)	0689	1.350	0.1000	2.576	2.103	1.089	16	-0.58	106.02	0
106	Vitamin A (KU / kg)	0004	1.410	0.2600	2.576	2.103	1.089	16	-0.55	106.02	0
106	Vitamin A (KU / kg)	0910	1.700	0.0000	2.576	2.103	1.089	16	-0.42	106.02	0
106	Vitamin A (KU / kg)	0010	1.840	0.5200	2.576	2.103	1.089	16	-0.35	106.02	0
106	Vitamin A (KU / kg)	0036	2.600	1.200	2.576	2.103	1.089	16	0.01	106.02	0
106	Vitamin A (KU / kg)	0038	2.660	0.4000	2.576	2.103	1.089	16	0.04	106.02	0
106	Vitamin A (KU / kg)	0098	3.715	4.750	2.576	2.103	1.089	16	0.54	106.01	0
106	Vitamin A (KU / kg)	0619	4.125	0.1900	2.576	2.103	1.089	16	0.74	106.02	0
106	Vitamin A (KU / kg)	0027	7.696	2.247	2.576	2.103	1.089	16	2.43	106.02	0
106	Vitamin A (KU / kg)	0675	12.03	3.610	2.576	2.103	1.089	16	4.49	106.02	0
106	Vitamin A (KU / kg)	0914	15.83	3.867	2.576	2.103	1.089	16	6.30	106.02	0
106	Vitamin A (KU / kg)	2319	< 4		2.576	2.103	1.089	16		106.02	5
107	Vitamin B12 (ppb)	0227	12.20	0.8000				2		107.00	0
107	Vitamin B12 (ppb)	0160	40.30	8.000				2		107.99	0
108	Vitamin D3 (KU / kg)	0619	0.2950	0.0100				3		108.99	0
108	Vitamin D3 (KU / kg)	0227	0.3655	0.0010				3		108.02	0
108	Vitamin D3 (KU / kg)	0010	0.8720	0.2180				3		108.99	0
108	Vitamin D3 (KU / kg)	0675	2.140	0.3400				3		108.02	2
108	Vitamin D3 (KU / kg)	2319	< 0.4					3		108.02	5
108	Vitamin D3 (KU / kg)	0910	< 0.5					3		108.02	5
108	Vitamin D3 (KU / kg)	0904	< 0.5					3		108.99	5
109	Vitamin E (IU / kg)	0619	31.00	2.000	41.45	7.829	4.001	13	-1.34	109.02	0
109	Vitamin E (IU / kg)	0905	34.50	3.000	41.45	7.829	4.001	13	-0.89	109.02	0
109	Vitamin E (IU / kg)	0904	34.62	5.730	41.45	7.829	4.001	13	-0.87	109.99	0
109	Vitamin E (IU / kg)	0910	37.10	2.200	41.45	7.829	4.001	13	-0.56	109.02	0
109	Vitamin E (IU / kg)	0171	38.00	10.00	41.45	7.829	4.001	13	-0.44	109.99	0
109	Vitamin E (IU / kg)	0676	38.98	0.1900	41.45	7.829	4.001	13	-0.32	109.02	0
109	Vitamin E (IU / kg)	2103	39.00	6.000	41.45	7.829	4.001	13	-0.31	109.02	0
109	Vitamin E (IU / kg)	0227	43.45	3.100	41.45	7.829	4.001	13	0.26	109.02	0
109	Vitamin E (IU / kg)	0036	44.65	0.9000	41.45	7.829	4.001	13	0.41	109.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
109	Vitamin E (IU / kg)	0148	45.30	3.700	41.45	7.829	4.001	13	0.49	109.02	0
109	Vitamin E (IU / kg)	0098	45.90	13.80	41.45	7.829	4.001	13	0.57	109.02	0
109	Vitamin E (IU / kg)	0563	54.17	0.3799	41.45	7.829	4.001	13	1.62	109.02	0
109	Vitamin E (IU / kg)	0675	85.24	1.010	41.45	7.829	4.001	13	5.59	109.02	0
111	Vitamin C, Phosphorylated (ppm)	0227	< 0.044					0		111.00	5
112	Pyridoxine (µg / g)	0227	2.690	0.0400				1		112.01	0
113	Folic Acid (ppm)	0227	0.6835	0.0670				1		113.01	0
114	Biotin (ppm)	0227	0.2460	0.0100				1		114.01	0
118	Peroxide value (meq/kg)	2081	7.555	4.590				1		118.99	0
120	Alanine (%)	0889	0.7100	0.0200	0.7984	0.0247	0.0116	29	-3.58	120.99	0
120	Alanine (%)	0353	0.7150	0.0100	0.7984	0.0247	0.0116	29	-3.38	120.00	0
120	Alanine (%)	0652	0.7550	0.0500	0.7984	0.0247	0.0116	29	-1.76	120.00	0
120	Alanine (%)	0910	0.7750	0.0100	0.7984	0.0247	0.0116	29	-0.95	120.00	0
120	Alanine (%)	0148	0.7750	0.0100	0.7984	0.0247	0.0116	29	-0.95	120.05	0
120	Alanine (%)	0354	0.7755	0.0010	0.7984	0.0247	0.0116	29	-0.93	120.00	0
120	Alanine (%)	0859	0.7803	0.0162	0.7984	0.0247	0.0116	29	-0.73	120.00	0
120	Alanine (%)	0619	0.7815	0.0150	0.7984	0.0247	0.0116	29	-0.68	120.00	0
120	Alanine (%)	0675	0.7850	0.0300	0.7984	0.0247	0.0116	29	-0.54	120.00	0
120	Alanine (%)	2146	0.7850	0.0100	0.7984	0.0247	0.0116	29	-0.54	120.99	0
120	Alanine (%)	0968	0.7950	0.0100	0.7984	0.0247	0.0116	29	-0.14	120.00	0
120	Alanine (%)	0626	0.7950	0.0020	0.7984	0.0247	0.0116	29	-0.14	120.05	0
120	Alanine (%)	0676	0.7950	0.0100	0.7984	0.0247	0.0116	29	-0.14	120.05	0
120	Alanine (%)	0918	0.7980	0.0495	0.7984	0.0247	0.0116	29	-0.02	120.00	0
120	Alanine (%)	0682	0.7990	0.0000	0.7984	0.0247	0.0116	29	0.03	120.00	0
120	Alanine (%)	0941	0.7995	0.0110	0.7984	0.0247	0.0116	29	0.05	120.00	0
120	Alanine (%)	0872	0.8025	0.0070	0.7984	0.0247	0.0116	29	0.17	120.00	0
120	Alanine (%)	0878	0.8045	0.0030	0.7984	0.0247	0.0116	29	0.25	120.00	0
120	Alanine (%)	0644	0.8070	0.0020	0.7984	0.0247	0.0116	29	0.35	120.00	0
120	Alanine (%)	0868	0.8080	0.0180	0.7984	0.0247	0.0116	29	0.39	120.00	0
120	Alanine (%)	0098	0.8115	0.0090	0.7984	0.0247	0.0116	29	0.53	120.02	0
120	Alanine (%)	0407	0.8130	0.0180	0.7984	0.0247	0.0116	29	0.59	120.05	0
120	Alanine (%)	2059	0.8165	0.0050	0.7984	0.0247	0.0116	29	0.73	120.00	0
120	Alanine (%)	0227	0.8200	0.0000	0.7984	0.0247	0.0116	29	0.88	120.99	0
120	Alanine (%)	0571	0.8205	0.0150	0.7984	0.0247	0.0116	29	0.90	120.00	0
120	Alanine (%)	0870	0.8205	0.0022	0.7984	0.0247	0.0116	29	0.90	120.00	0
120	Alanine (%)	0226	0.8359	0.0021	0.7984	0.0247	0.0116	29	1.52	120.00	0
120	Alanine (%)	2188	0.8490	0.0000	0.7984	0.0247	0.0116	29	2.05	120.05	0
120	Alanine (%)	2196	0.8660	0.0000	0.7984	0.0247	0.0116	29	2.74	120.05	0
121	Arginine (%)	0652	0.9150	0.0100	0.9948	0.0403	0.0157	27	-1.98	121.00	0
121	Arginine (%)	0859	0.9341	0.0188	0.9948	0.0403	0.0157	27	-1.51	121.00	0
121	Arginine (%)	0407	0.9375	0.0190	0.9948	0.0403	0.0157	27	-1.42	121.05	0
121	Arginine (%)	0682	0.9510	0.0000	0.9948	0.0403	0.0157	27	-1.09	121.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
121	Arginine (%)	0889	0.9550	0.0300	0.9948	0.0403	0.0157	27	-0.99	121.99	0
121	Arginine (%)	0968	0.9600	0.0000	0.9948	0.0403	0.0157	27	-0.86	121.00	0
121	Arginine (%)	0918	0.9604	0.0579	0.9948	0.0403	0.0157	27	-0.85	121.00	0
121	Arginine (%)	0619	0.9695	0.0070	0.9948	0.0403	0.0157	27	-0.63	121.00	0
121	Arginine (%)	0675	0.9850	0.0300	0.9948	0.0403	0.0157	27	-0.24	121.00	0
121	Arginine (%)	0148	0.9850	0.0700	0.9948	0.0403	0.0157	27	-0.24	121.05	0
121	Arginine (%)	0868	0.9900	0.0220	0.9948	0.0403	0.0157	27	-0.12	121.00	0
121	Arginine (%)	0626	0.9900	0.0040	0.9948	0.0403	0.0157	27	-0.12	121.05	0
121	Arginine (%)	0870	0.9914	0.0122	0.9948	0.0403	0.0157	27	-0.08	121.00	0
121	Arginine (%)	0353	0.9950	0.0100	0.9948	0.0403	0.0157	27	0.01	121.00	0
121	Arginine (%)	0910	1.000	0.0000	0.9948	0.0403	0.0157	27	0.13	121.00	0
121	Arginine (%)	0227	1.000	0.0200	0.9948	0.0403	0.0157	27	0.13	121.99	0
121	Arginine (%)	0941	1.015	0.0090	0.9948	0.0403	0.0157	27	0.49	121.00	0
121	Arginine (%)	0098	1.015	0.0200	0.9948	0.0403	0.0157	27	0.50	121.02	0
121	Arginine (%)	2188	1.015	0.0000	0.9948	0.0403	0.0157	27	0.50	121.05	0
121	Arginine (%)	0878	1.020	0.0050	0.9948	0.0403	0.0157	27	0.61	121.00	0
121	Arginine (%)	0354	1.021	0.0110	0.9948	0.0403	0.0157	27	0.64	121.00	0
121	Arginine (%)	0644	1.026	0.0050	0.9948	0.0403	0.0157	27	0.76	121.00	0
121	Arginine (%)	0872	1.026	0.0090	0.9948	0.0403	0.0157	27	0.76	121.00	0
121	Arginine (%)	0571	1.030	0.0280	0.9948	0.0403	0.0157	27	0.87	121.00	0
121	Arginine (%)	2059	1.045	0.0080	0.9948	0.0403	0.0157	27	1.25	121.00	0
121	Arginine (%)	2196	1.112	0.0000	0.9948	0.0403	0.0157	27	2.91	121.05	0
121	Arginine (%)	0226	1.185	0.0169	0.9948	0.0403	0.0157	27	4.72	121.00	0
121	Arginine (%)	0676	1.070	0.1400	0.9948	0.0403	0.0157	27	1.87	121.05	1
122	Aspartic (%)	0889	1.215	0.0100	1.400	0.0425	0.0224	28	-4.35	122.99	0
122	Aspartic (%)	0652	1.275	0.0300	1.400	0.0425	0.0224	28	-2.94	122.00	0
122	Aspartic (%)	0859	1.309	0.0230	1.400	0.0425	0.0224	28	-2.14	122.00	0
122	Aspartic (%)	0675	1.345	0.0100	1.400	0.0425	0.0224	28	-1.29	122.00	0
122	Aspartic (%)	0619	1.355	0.0300	1.400	0.0425	0.0224	28	-1.06	122.00	0
122	Aspartic (%)	0968	1.355	0.0100	1.400	0.0425	0.0224	28	-1.06	122.00	0
122	Aspartic (%)	0682	1.372	0.0000	1.400	0.0425	0.0224	28	-0.66	122.00	0
122	Aspartic (%)	0941	1.376	0.0520	1.400	0.0425	0.0224	28	-0.57	122.00	0
122	Aspartic (%)	0354	1.383	0.0150	1.400	0.0425	0.0224	28	-0.41	122.00	0
122	Aspartic (%)	0353	1.395	0.0100	1.400	0.0425	0.0224	28	-0.12	122.00	0
122	Aspartic (%)	2188	1.396	0.0000	1.400	0.0425	0.0224	28	-0.09	122.05	0
122	Aspartic (%)	0918	1.396	0.0895	1.400	0.0425	0.0224	28	-0.09	122.00	0
122	Aspartic (%)	0644	1.400	0.0090	1.400	0.0425	0.0224	28	-0.01	122.00	0
122	Aspartic (%)	0910	1.405	0.0300	1.400	0.0425	0.0224	28	0.12	122.00	0
122	Aspartic (%)	0148	1.405	0.0300	1.400	0.0425	0.0224	28	0.12	122.05	0
122	Aspartic (%)	2146	1.405	0.0100	1.400	0.0425	0.0224	28	0.12	122.99	0
122	Aspartic (%)	0872	1.408	0.0120	1.400	0.0425	0.0224	28	0.19	122.00	0
122	Aspartic (%)	0878	1.416	0.0020	1.400	0.0425	0.0224	28	0.38	122.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
122	Aspartic (%)	0571	1.418	0.0200	1.400	0.0425	0.0224	28	0.42	122.00	0
122	Aspartic (%)	2059	1.422	0.0140	1.400	0.0425	0.0224	28	0.52	122.00	0
122	Aspartic (%)	0870	1.430	0.0410	1.400	0.0425	0.0224	28	0.69	122.00	0
122	Aspartic (%)	0098	1.434	0.0480	1.400	0.0425	0.0224	28	0.80	122.02	0
122	Aspartic (%)	0676	1.435	0.0500	1.400	0.0425	0.0224	28	0.82	122.05	0
122	Aspartic (%)	0868	1.436	0.0230	1.400	0.0425	0.0224	28	0.84	122.00	0
122	Aspartic (%)	0626	1.439	0.0030	1.400	0.0425	0.0224	28	0.91	122.05	0
122	Aspartic (%)	0227	1.440	0.0200	1.400	0.0425	0.0224	28	0.94	122.99	0
122	Aspartic (%)	2196	1.518	0.0000	1.400	0.0425	0.0224	28	2.78	122.05	0
122	Aspartic (%)	0226	1.548	0.0363	1.400	0.0425	0.0224	28	3.48	122.00	0
122	Aspartic (%)	0407	1.624	0.2850	1.400	0.0425	0.0224	28	5.26	122.05	1
124	Cysteine/Cystine (%)	0870	0.1951	0.0130	0.2589	0.0220	0.0094	28	-2.91	124.00	0
124	Cysteine/Cystine (%)	0407	0.2083	0.0258	0.2589	0.0220	0.0094	28	-2.30	124.05	0
124	Cysteine/Cystine (%)	0859	0.2296	0.0044	0.2589	0.0220	0.0094	28	-1.33	124.00	0
124	Cysteine/Cystine (%)	0227	0.2350	0.0100	0.2589	0.0220	0.0094	28	-1.09	124.99	0
124	Cysteine/Cystine (%)	0652	0.2400	0.0200	0.2589	0.0220	0.0094	28	-0.86	124.00	0
124	Cysteine/Cystine (%)	0889	0.2400	0.0200	0.2589	0.0220	0.0094	28	-0.86	124.99	0
124	Cysteine/Cystine (%)	2146	0.2400	0.0000	0.2589	0.0220	0.0094	28	-0.86	124.99	0
124	Cysteine/Cystine (%)	0226	0.2430	0.0177	0.2589	0.0220	0.0094	28	-0.73	124.00	0
124	Cysteine/Cystine (%)	0910	0.2500	0.0200	0.2589	0.0220	0.0094	28	-0.41	124.00	0
124	Cysteine/Cystine (%)	2196	0.2510	0.0000	0.2589	0.0220	0.0094	28	-0.36	124.05	0
124	Cysteine/Cystine (%)	0619	0.2550	0.0000	0.2589	0.0220	0.0094	28	-0.18	124.00	0
124	Cysteine/Cystine (%)	0676	0.2550	0.0100	0.2589	0.0220	0.0094	28	-0.18	124.05	0
124	Cysteine/Cystine (%)	0878	0.2565	0.0010	0.2589	0.0220	0.0094	28	-0.11	124.00	0
124	Cysteine/Cystine (%)	0872	0.2610	0.0020	0.2589	0.0220	0.0094	28	0.10	124.00	0
124	Cysteine/Cystine (%)	0148	0.2615	0.0270	0.2589	0.0220	0.0094	28	0.12	124.05	0
124	Cysteine/Cystine (%)	0941	0.2620	0.0180	0.2589	0.0220	0.0094	28	0.14	124.00	0
124	Cysteine/Cystine (%)	2188	0.2640	0.0000	0.2589	0.0220	0.0094	28	0.23	124.05	0
124	Cysteine/Cystine (%)	0353	0.2650	0.0100	0.2589	0.0220	0.0094	28	0.28	124.00	0
124	Cysteine/Cystine (%)	0644	0.2675	0.0190	0.2589	0.0220	0.0094	28	0.39	124.00	0
124	Cysteine/Cystine (%)	2059	0.2675	0.0010	0.2589	0.0220	0.0094	28	0.39	124.00	0
124	Cysteine/Cystine (%)	0571	0.2690	0.0060	0.2589	0.0220	0.0094	28	0.46	124.00	0
124	Cysteine/Cystine (%)	0968	0.2700	0.0000	0.2589	0.0220	0.0094	28	0.51	124.00	0
124	Cysteine/Cystine (%)	0918	0.2750	0.0100	0.2589	0.0220	0.0094	28	0.73	124.00	0
124	Cysteine/Cystine (%)	0868	0.2805	0.0070	0.2589	0.0220	0.0094	28	0.98	124.00	0
124	Cysteine/Cystine (%)	0675	0.2850	0.0100	0.2589	0.0220	0.0094	28	1.19	124.00	0
124	Cysteine/Cystine (%)	0098	0.2905	0.0050	0.2589	0.0220	0.0094	28	1.44	124.02	0
124	Cysteine/Cystine (%)	0682	0.3140	0.0000	0.2589	0.0220	0.0094	28	2.51	124.00	0
124	Cysteine/Cystine (%)	0354	0.3195	0.0050	0.2589	0.0220	0.0094	28	2.76	124.00	0
124	Cysteine/Cystine (%)	2319	< 0.005		0.2589	0.0220	0.0094	28		124.05	5
125	Glutamic (%)	0889	2.365	0.0600	2.786	0.1174	0.0442	28	-3.59	125.99	0
125	Glutamic (%)	2188	2.568	0.0000	2.786	0.1174	0.0442	28	-1.86	125.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
125	Glutamic (%)	0859	2.642	0.0514	2.786	0.1174	0.0442	28	-1.22	125.00	0
125	Glutamic (%)	0652	2.670	0.0200	2.786	0.1174	0.0442	28	-0.99	125.00	0
125	Glutamic (%)	0682	2.681	0.0000	2.786	0.1174	0.0442	28	-0.89	125.00	0
125	Glutamic (%)	0968	2.710	0.0200	2.786	0.1174	0.0442	28	-0.65	125.00	0
125	Glutamic (%)	0675	2.720	0.0800	2.786	0.1174	0.0442	28	-0.56	125.00	0
125	Glutamic (%)	0941	2.726	0.0440	2.786	0.1174	0.0442	28	-0.51	125.00	0
125	Glutamic (%)	0354	2.728	0.0140	2.786	0.1174	0.0442	28	-0.49	125.00	0
125	Glutamic (%)	0868	2.738	0.0370	2.786	0.1174	0.0442	28	-0.41	125.00	0
125	Glutamic (%)	0910	2.745	0.0700	2.786	0.1174	0.0442	28	-0.35	125.00	0
125	Glutamic (%)	0571	2.756	0.0240	2.786	0.1174	0.0442	28	-0.25	125.00	0
125	Glutamic (%)	0918	2.760	0.1935	2.786	0.1174	0.0442	28	-0.22	125.00	0
125	Glutamic (%)	0872	2.764	0.0190	2.786	0.1174	0.0442	28	-0.19	125.00	0
125	Glutamic (%)	2059	2.771	0.0150	2.786	0.1174	0.0442	28	-0.13	125.00	0
125	Glutamic (%)	0227	2.780	0.0000	2.786	0.1174	0.0442	28	-0.05	125.99	0
125	Glutamic (%)	0644	2.792	0.0120	2.786	0.1174	0.0442	28	0.05	125.00	0
125	Glutamic (%)	0619	2.810	0.0400	2.786	0.1174	0.0442	28	0.21	125.00	0
125	Glutamic (%)	0098	2.824	0.0480	2.786	0.1174	0.0442	28	0.32	125.02	0
125	Glutamic (%)	0353	2.825	0.0100	2.786	0.1174	0.0442	28	0.33	125.00	0
125	Glutamic (%)	0407	2.842	0.1800	2.786	0.1174	0.0442	28	0.48	125.05	0
125	Glutamic (%)	0878	2.849	0.0080	2.786	0.1174	0.0442	28	0.54	125.00	0
125	Glutamic (%)	0626	2.915	0.0080	2.786	0.1174	0.0442	28	1.10	125.05	0
125	Glutamic (%)	0870	2.927	0.0732	2.786	0.1174	0.0442	28	1.20	125.00	0
125	Glutamic (%)	2196	2.949	0.0000	2.786	0.1174	0.0442	28	1.39	125.05	0
125	Glutamic (%)	0148	2.950	0.1000	2.786	0.1174	0.0442	28	1.40	125.05	0
125	Glutamic (%)	2146	2.950	0.0400	2.786	0.1174	0.0442	28	1.40	125.99	0
125	Glutamic (%)	0226	3.011	0.0700	2.786	0.1174	0.0442	28	1.92	125.00	0
125	Glutamic (%)	0676	3.310	0.2600	2.786	0.1174	0.0442	28	4.47	125.05	1
126	Glycine (%)	0676	0.6400	0.0000	0.7510	0.0246	0.0112	27	-4.51	126.05	0
126	Glycine (%)	0652	0.7150	0.0300	0.7510	0.0246	0.0112	27	-1.46	126.00	0
126	Glycine (%)	0859	0.7162	0.0142	0.7510	0.0246	0.0112	27	-1.41	126.00	0
126	Glycine (%)	0682	0.7190	0.0000	0.7510	0.0246	0.0112	27	-1.30	126.00	0
126	Glycine (%)	0353	0.7300	0.0000	0.7510	0.0246	0.0112	27	-0.85	126.00	0
126	Glycine (%)	0910	0.7300	0.0000	0.7510	0.0246	0.0112	27	-0.85	126.00	0
126	Glycine (%)	0675	0.7350	0.0300	0.7510	0.0246	0.0112	27	-0.65	126.00	0
126	Glycine (%)	2146	0.7350	0.0100	0.7510	0.0246	0.0112	27	-0.65	126.99	0
126	Glycine (%)	2188	0.7420	0.0000	0.7510	0.0246	0.0112	27	-0.37	126.05	0
126	Glycine (%)	0968	0.7450	0.0100	0.7510	0.0246	0.0112	27	-0.24	126.00	0
126	Glycine (%)	0878	0.7500	0.0020	0.7510	0.0246	0.0112	27	-0.04	126.00	0
126	Glycine (%)	0148	0.7500	0.0200	0.7510	0.0246	0.0112	27	-0.04	126.05	0
126	Glycine (%)	0354	0.7520	0.0100	0.7510	0.0246	0.0112	27	0.04	126.00	0
126	Glycine (%)	0870	0.7529	0.0023	0.7510	0.0246	0.0112	27	0.08	126.00	0
126	Glycine (%)	0941	0.7530	0.0080	0.7510	0.0246	0.0112	27	0.08	126.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
126	Glycine (%)	0918	0.7570	0.0504	0.7510	0.0246	0.0112	27	0.24	126.00	0
126	Glycine (%)	0619	0.7575	0.0150	0.7510	0.0246	0.0112	27	0.26	126.00	0
126	Glycine (%)	0872	0.7580	0.0040	0.7510	0.0246	0.0112	27	0.28	126.00	0
126	Glycine (%)	0868	0.7620	0.0160	0.7510	0.0246	0.0112	27	0.45	126.00	0
126	Glycine (%)	0571	0.7625	0.0070	0.7510	0.0246	0.0112	27	0.47	126.00	0
126	Glycine (%)	0644	0.7645	0.0010	0.7510	0.0246	0.0112	27	0.55	126.00	0
126	Glycine (%)	2059	0.7700	0.0080	0.7510	0.0246	0.0112	27	0.77	126.00	0
126	Glycine (%)	0227	0.7700	0.0200	0.7510	0.0246	0.0112	27	0.77	126.99	0
126	Glycine (%)	0098	0.7725	0.0110	0.7510	0.0246	0.0112	27	0.87	126.02	0
126	Glycine (%)	2196	0.8030	0.0000	0.7510	0.0246	0.0112	27	2.11	126.05	0
126	Glycine (%)	0626	0.8060	0.0040	0.7510	0.0246	0.0112	27	2.23	126.05	0
126	Glycine (%)	0226	0.8443	0.0291	0.7510	0.0246	0.0112	27	3.79	126.00	0
126	Glycine (%)	0407	0.7665	0.1710	0.7510	0.0246	0.0112	27	0.63	126.05	1
126	Glycine (%)	0889	0.4025	0.0050	0.7510	0.0246	0.0112	27	-14.16	126.99	2
127	Histidine (%)	2188	0.3370	0.0000	0.3919	0.0222	0.0094	28	-2.48	127.05	0
127	Histidine (%)	0626	0.3600	0.0100	0.3919	0.0222	0.0094	28	-1.44	127.05	0
127	Histidine (%)	0870	0.3635	0.0194	0.3919	0.0222	0.0094	28	-1.28	127.00	0
127	Histidine (%)	0859	0.3636	0.0091	0.3919	0.0222	0.0094	28	-1.28	127.00	0
127	Histidine (%)	2146	0.3650	0.0100	0.3919	0.0222	0.0094	28	-1.21	127.99	0
127	Histidine (%)	0353	0.3750	0.0100	0.3919	0.0222	0.0094	28	-0.76	127.00	0
127	Histidine (%)	0676	0.3750	0.0500	0.3919	0.0222	0.0094	28	-0.76	127.05	0
127	Histidine (%)	0675	0.3800	0.0000	0.3919	0.0222	0.0094	28	-0.54	127.00	0
127	Histidine (%)	0968	0.3800	0.0000	0.3919	0.0222	0.0094	28	-0.54	127.00	0
127	Histidine (%)	0227	0.3850	0.0100	0.3919	0.0222	0.0094	28	-0.31	127.99	0
127	Histidine (%)	0652	0.3900	0.0200	0.3919	0.0222	0.0094	28	-0.09	127.00	0
127	Histidine (%)	0148	0.3900	0.0200	0.3919	0.0222	0.0094	28	-0.09	127.05	0
127	Histidine (%)	0868	0.3910	0.0060	0.3919	0.0222	0.0094	28	-0.04	127.00	0
127	Histidine (%)	0571	0.3945	0.0110	0.3919	0.0222	0.0094	28	0.12	127.00	0
127	Histidine (%)	0910	0.3950	0.0100	0.3919	0.0222	0.0094	28	0.14	127.00	0
127	Histidine (%)	0354	0.3970	0.0120	0.3919	0.0222	0.0094	28	0.23	127.00	0
127	Histidine (%)	0872	0.3985	0.0070	0.3919	0.0222	0.0094	28	0.30	127.00	0
127	Histidine (%)	0098	0.3995	0.0070	0.3919	0.0222	0.0094	28	0.34	127.02	0
127	Histidine (%)	2059	0.4005	0.0030	0.3919	0.0222	0.0094	28	0.39	127.00	0
127	Histidine (%)	0644	0.4015	0.0050	0.3919	0.0222	0.0094	28	0.43	127.00	0
127	Histidine (%)	0878	0.4025	0.0010	0.3919	0.0222	0.0094	28	0.48	127.00	0
127	Histidine (%)	0889	0.4025	0.0050	0.3919	0.0222	0.0094	28	0.48	127.99	0
127	Histidine (%)	0941	0.4030	0.0080	0.3919	0.0222	0.0094	28	0.50	127.00	0
127	Histidine (%)	0682	0.4070	0.0000	0.3919	0.0222	0.0094	28	0.68	127.00	0
127	Histidine (%)	0918	0.4193	0.0212	0.3919	0.0222	0.0094	28	1.24	127.00	0
127	Histidine (%)	0619	0.4360	0.0020	0.3919	0.0222	0.0094	28	1.99	127.00	0
127	Histidine (%)	0226	0.4598	0.0065	0.3919	0.0222	0.0094	28	3.06	127.00	0
127	Histidine (%)	2196	0.4600	0.0000	0.3919	0.0222	0.0094	28	3.07	127.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
127	Histidine (%)	0407	0.3415	0.0530	0.3919	0.0222	0.0094	28	-2.27	127.05	1
128	Isoleucine (%)	0353	0.4650	0.0100	0.5742	0.0318	0.0118	28	-3.43	128.00	0
128	Isoleucine (%)	0859	0.5259	0.0121	0.5742	0.0318	0.0118	28	-1.52	128.00	0
128	Isoleucine (%)	0889	0.5300	0.0000	0.5742	0.0318	0.0118	28	-1.39	128.99	0
128	Isoleucine (%)	0682	0.5320	0.0000	0.5742	0.0318	0.0118	28	-1.33	128.00	0
128	Isoleucine (%)	0626	0.5380	0.0000	0.5742	0.0318	0.0118	28	-1.14	128.05	0
128	Isoleucine (%)	0918	0.5445	0.0184	0.5742	0.0318	0.0118	28	-0.93	128.00	0
128	Isoleucine (%)	0354	0.5515	0.0450	0.5742	0.0318	0.0118	28	-0.71	128.00	0
128	Isoleucine (%)	0675	0.5550	0.0100	0.5742	0.0318	0.0118	28	-0.60	128.00	0
128	Isoleucine (%)	0968	0.5550	0.0100	0.5742	0.0318	0.0118	28	-0.60	128.00	0
128	Isoleucine (%)	0652	0.5650	0.0100	0.5742	0.0318	0.0118	28	-0.29	128.00	0
128	Isoleucine (%)	0870	0.5724	0.0133	0.5742	0.0318	0.0118	28	-0.06	128.00	0
128	Isoleucine (%)	2188	0.5760	0.0000	0.5742	0.0318	0.0118	28	0.06	128.05	0
128	Isoleucine (%)	0868	0.5775	0.0050	0.5742	0.0318	0.0118	28	0.10	128.00	0
128	Isoleucine (%)	0910	0.5800	0.0000	0.5742	0.0318	0.0118	28	0.18	128.00	0
128	Isoleucine (%)	0676	0.5800	0.0400	0.5742	0.0318	0.0118	28	0.18	128.05	0
128	Isoleucine (%)	0226	0.5827	0.0636	0.5742	0.0318	0.0118	28	0.27	128.00	0
128	Isoleucine (%)	0619	0.5840	0.0060	0.5742	0.0318	0.0118	28	0.31	128.00	0
128	Isoleucine (%)	0878	0.5845	0.0030	0.5742	0.0318	0.0118	28	0.32	128.00	0
128	Isoleucine (%)	2146	0.5850	0.0100	0.5742	0.0318	0.0118	28	0.34	128.99	0
128	Isoleucine (%)	0872	0.5855	0.0190	0.5742	0.0318	0.0118	28	0.35	128.00	0
128	Isoleucine (%)	0227	0.5950	0.0100	0.5742	0.0318	0.0118	28	0.65	128.99	0
128	Isoleucine (%)	0098	0.5960	0.0060	0.5742	0.0318	0.0118	28	0.69	128.02	0
128	Isoleucine (%)	2059	0.5985	0.0010	0.5742	0.0318	0.0118	28	0.76	128.00	0
128	Isoleucine (%)	0571	0.6010	0.0100	0.5742	0.0318	0.0118	28	0.84	128.00	0
128	Isoleucine (%)	0941	0.6010	0.0060	0.5742	0.0318	0.0118	28	0.84	128.00	0
128	Isoleucine (%)	0644	0.6130	0.0020	0.5742	0.0318	0.0118	28	1.22	128.00	0
128	Isoleucine (%)	0148	0.6200	0.0200	0.5742	0.0318	0.0118	28	1.44	128.05	0
128	Isoleucine (%)	2196	0.6960	0.0000	0.5742	0.0318	0.0118	28	3.83	128.05	0
128	Isoleucine (%)	0407	0.5200	0.1700	0.5742	0.0318	0.0118	28	-1.70	128.05	1
128	Isoleucine (%)	2319	< 0.005		0.5742	0.0318	0.0118	28		128.05	5
129	Leucine (%)	0889	1.048	0.0350	1.214	0.0443	0.0158	28	-3.77	129.99	0
129	Leucine (%)	0652	1.125	0.0100	1.214	0.0443	0.0158	28	-2.02	129.00	0
129	Leucine (%)	0682	1.125	0.0000	1.214	0.0443	0.0158	28	-2.02	129.00	0
129	Leucine (%)	2188	1.138	0.0000	1.214	0.0443	0.0158	28	-1.73	129.05	0
129	Leucine (%)	0859	1.158	0.0245	1.214	0.0443	0.0158	28	-1.27	129.00	0
129	Leucine (%)	0626	1.183	0.0100	1.214	0.0443	0.0158	28	-0.71	129.05	0
129	Leucine (%)	2146	1.185	0.0100	1.214	0.0443	0.0158	28	-0.67	129.99	0
129	Leucine (%)	0910	1.190	0.0200	1.214	0.0443	0.0158	28	-0.55	129.00	0
129	Leucine (%)	0675	1.200	0.0400	1.214	0.0443	0.0158	28	-0.33	129.00	0
129	Leucine (%)	0968	1.200	0.0000	1.214	0.0443	0.0158	28	-0.33	129.00	0
129	Leucine (%)	0870	1.204	0.0140	1.214	0.0443	0.0158	28	-0.23	129.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
129	Leucine (%)	0354	1.216	0.0000	1.214	0.0443	0.0158	28	0.03	129.00	0
129	Leucine (%)	0148	1.220	0.0400	1.214	0.0443	0.0158	28	0.12	129.05	0
129	Leucine (%)	0941	1.220	0.0180	1.214	0.0443	0.0158	28	0.12	129.00	0
129	Leucine (%)	0878	1.223	0.0060	1.214	0.0443	0.0158	28	0.19	129.00	0
129	Leucine (%)	0353	1.225	0.0100	1.214	0.0443	0.0158	28	0.24	129.00	0
129	Leucine (%)	0676	1.230	0.0600	1.214	0.0443	0.0158	28	0.35	129.05	0
129	Leucine (%)	0571	1.237	0.0090	1.214	0.0443	0.0158	28	0.50	129.00	0
129	Leucine (%)	0644	1.242	0.0050	1.214	0.0443	0.0158	28	0.61	129.00	0
129	Leucine (%)	0868	1.243	0.0110	1.214	0.0443	0.0158	28	0.63	129.00	0
129	Leucine (%)	2059	1.243	0.0050	1.214	0.0443	0.0158	28	0.63	129.00	0
129	Leucine (%)	0098	1.243	0.0160	1.214	0.0443	0.0158	28	0.64	129.02	0
129	Leucine (%)	0872	1.244	0.0050	1.214	0.0443	0.0158	28	0.66	129.00	0
129	Leucine (%)	0619	1.245	0.0100	1.214	0.0443	0.0158	28	0.69	129.00	0
129	Leucine (%)	0227	1.250	0.0000	1.214	0.0443	0.0158	28	0.80	129.99	0
129	Leucine (%)	0918	1.252	0.0585	1.214	0.0443	0.0158	28	0.86	129.00	0
129	Leucine (%)	0226	1.288	0.0249	1.214	0.0443	0.0158	28	1.66	129.00	0
129	Leucine (%)	2196	1.319	0.0000	1.214	0.0443	0.0158	28	2.36	129.05	0
129	Leucine (%)	0407	1.107	0.3050	1.214	0.0443	0.0158	28	-2.44	129.05	1
129	Leucine (%)	2319	< 0.005		1.214	0.0443	0.0158	28		129.05	5
130	L-Lysine (%)	2144	0.6400	0.0000	0.7323	0.0392	0.0159	32	-2.35	130.99	0
130	L-Lysine (%)	0652	0.6750	0.0500	0.7323	0.0392	0.0159	32	-1.46	130.00	0
130	L-Lysine (%)	0859	0.6840	0.0147	0.7323	0.0392	0.0159	32	-1.23	130.00	0
130	L-Lysine (%)	0848	0.6900	0.0000	0.7323	0.0392	0.0159	32	-1.08	130.00	0
130	L-Lysine (%)	0918	0.6910	0.0000	0.7323	0.0392	0.0159	32	-1.05	130.00	0
130	L-Lysine (%)	2146	0.6950	0.0100	0.7323	0.0392	0.0159	32	-0.95	130.99	0
130	L-Lysine (%)	0868	0.7000	0.0120	0.7323	0.0392	0.0159	32	-0.82	130.00	0
130	L-Lysine (%)	2188	0.7020	0.0000	0.7323	0.0392	0.0159	32	-0.77	130.05	0
130	L-Lysine (%)	0619	0.7045	0.0550	0.7323	0.0392	0.0159	32	-0.71	130.00	0
130	L-Lysine (%)	0682	0.7050	0.0000	0.7323	0.0392	0.0159	32	-0.70	130.00	0
130	L-Lysine (%)	2053	0.7100	0.0000	0.7323	0.0392	0.0159	32	-0.57	130.99	0
130	L-Lysine (%)	0353	0.7150	0.0100	0.7323	0.0392	0.0159	32	-0.44	130.00	0
130	L-Lysine (%)	0941	0.7200	0.0140	0.7323	0.0392	0.0159	32	-0.31	130.00	0
130	L-Lysine (%)	0872	0.7235	0.0090	0.7323	0.0392	0.0159	32	-0.22	130.00	0
130	L-Lysine (%)	0675	0.7250	0.0100	0.7323	0.0392	0.0159	32	-0.19	130.00	0
130	L-Lysine (%)	0644	0.7345	0.0010	0.7323	0.0392	0.0159	32	0.06	130.00	0
130	L-Lysine (%)	0878	0.7345	0.0010	0.7323	0.0392	0.0159	32	0.06	130.00	0
130	L-Lysine (%)	0968	0.7350	0.0100	0.7323	0.0392	0.0159	32	0.07	130.00	0
130	L-Lysine (%)	0571	0.7475	0.0150	0.7323	0.0392	0.0159	32	0.39	130.00	0
130	L-Lysine (%)	2059	0.7490	0.0060	0.7323	0.0392	0.0159	32	0.43	130.00	0
130	L-Lysine (%)	0354	0.7540	0.0200	0.7323	0.0392	0.0159	32	0.55	130.00	0
130	L-Lysine (%)	0870	0.7547	0.0092	0.7323	0.0392	0.0159	32	0.57	130.00	0
130	L-Lysine (%)	0676	0.7550	0.0700	0.7323	0.0392	0.0159	32	0.58	130.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT	Your	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests	Z Score	Method	
130	L-Lysine (%)	0014	0.7565	0.0110	0.7323	0.0392	0.0159	32	0.62	130.05	0
130	L-Lysine (%)	0226	0.7580	0.0027	0.7323	0.0392	0.0159	32	0.66	130.00	0
130	L-Lysine (%)	0148	0.7600	0.0400	0.7323	0.0392	0.0159	32	0.71	130.05	0
130	L-Lysine (%)	0910	0.7650	0.0500	0.7323	0.0392	0.0159	32	0.84	130.00	0
130	L-Lysine (%)	0227	0.7650	0.0100	0.7323	0.0392	0.0159	32	0.84	130.99	0
130	L-Lysine (%)	0098	0.7805	0.0270	0.7323	0.0392	0.0159	32	1.23	130.02	0
130	L-Lysine (%)	0626	0.7880	0.0100	0.7323	0.0392	0.0159	32	1.42	130.05	0
130	L-Lysine (%)	2196	0.8070	0.0000	0.7323	0.0392	0.0159	32	1.91	130.05	0
130	L-Lysine (%)	0889	0.8900	0.0400	0.7323	0.0392	0.0159	32	4.03	130.99	0
130	L-Lysine (%)	0407	0.8310	0.2120	0.7323	0.0392	0.0159	32	2.52	130.05	1
131	Methionine (%)	0652	0.1700	0.0000	0.2413	0.0184	0.0069	29	-3.88	131.00	0
131	Methionine (%)	0626	0.1995	0.0210	0.2413	0.0184	0.0069	29	-2.27	131.05	0
131	Methionine (%)	0148	0.2050	0.0100	0.2413	0.0184	0.0069	29	-1.97	131.05	0
131	Methionine (%)	0859	0.2199	0.0047	0.2413	0.0184	0.0069	29	-1.16	131.00	0
131	Methionine (%)	0682	0.2200	0.0000	0.2413	0.0184	0.0069	29	-1.16	131.00	0
131	Methionine (%)	2146	0.2200	0.0200	0.2413	0.0184	0.0069	29	-1.16	131.99	0
131	Methionine (%)	0675	0.2300	0.0200	0.2413	0.0184	0.0069	29	-0.61	131.00	0
131	Methionine (%)	0941	0.2320	0.0020	0.2413	0.0184	0.0069	29	-0.50	131.00	0
131	Methionine (%)	0918	0.2350	0.0100	0.2413	0.0184	0.0069	29	-0.34	131.00	0
131	Methionine (%)	0227	0.2350	0.0100	0.2413	0.0184	0.0069	29	-0.34	131.99	0
131	Methionine (%)	0868	0.2355	0.0030	0.2413	0.0184	0.0069	29	-0.31	131.00	0
131	Methionine (%)	0619	0.2395	0.0050	0.2413	0.0184	0.0069	29	-0.10	131.00	0
131	Methionine (%)	0098	0.2405	0.0030	0.2413	0.0184	0.0069	29	-0.04	131.02	0
131	Methionine (%)	0889	0.2425	0.0050	0.2413	0.0184	0.0069	29	0.07	131.99	0
131	Methionine (%)	0571	0.2430	0.0040	0.2413	0.0184	0.0069	29	0.09	131.00	0
131	Methionine (%)	0870	0.2442	0.0111	0.2413	0.0184	0.0069	29	0.16	131.00	0
131	Methionine (%)	2059	0.2450	0.0020	0.2413	0.0184	0.0069	29	0.20	131.00	0
131	Methionine (%)	0878	0.2455	0.0030	0.2413	0.0184	0.0069	29	0.23	131.00	0
131	Methionine (%)	0872	0.2495	0.0050	0.2413	0.0184	0.0069	29	0.45	131.00	0
131	Methionine (%)	0910	0.2500	0.0000	0.2413	0.0184	0.0069	29	0.48	131.00	0
131	Methionine (%)	0968	0.2500	0.0000	0.2413	0.0184	0.0069	29	0.48	131.00	0
131	Methionine (%)	0354	0.2505	0.0250	0.2413	0.0184	0.0069	29	0.50	131.00	0
131	Methionine (%)	0644	0.2540	0.0040	0.2413	0.0184	0.0069	29	0.69	131.00	0
131	Methionine (%)	0676	0.2550	0.0100	0.2413	0.0184	0.0069	29	0.75	131.05	0
131	Methionine (%)	2188	0.2560	0.0000	0.2413	0.0184	0.0069	29	0.80	131.05	0
131	Methionine (%)	2196	0.2590	0.0000	0.2413	0.0184	0.0069	29	0.96	131.05	0
131	Methionine (%)	0226	0.2664	0.0080	0.2413	0.0184	0.0069	29	1.37	131.00	0
131	Methionine (%)	0407	0.2857	0.0049	0.2413	0.0184	0.0069	29	2.41	131.05	0
131	Methionine (%)	0353	0.3050	0.0100	0.2413	0.0184	0.0069	29	3.47	131.00	0
131	Methionine (%)	2319	< 0.005		0.2413	0.0184	0.0069	29		131.05	5
132	Phenylalanine (%)	0682	0.6320	0.0000	0.7097	0.0367	0.0089	28	-2.12	132.00	0
132	Phenylalanine (%)	0619	0.6415	0.0150	0.7097	0.0367	0.0089	28	-1.86	132.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
132	Phenylalanine (%)	2188	0.6650	0.0000	0.7097	0.0367	0.0089	28	-1.22	132.05	0
132	Phenylalanine (%)	0968	0.6700	0.0000	0.7097	0.0367	0.0089	28	-1.08	132.00	0
132	Phenylalanine (%)	0626	0.6760	0.0000	0.7097	0.0367	0.0089	28	-0.92	132.05	0
132	Phenylalanine (%)	2146	0.6800	0.0000	0.7097	0.0367	0.0089	28	-0.81	132.99	0
132	Phenylalanine (%)	0675	0.6850	0.0100	0.7097	0.0367	0.0089	28	-0.67	132.00	0
132	Phenylalanine (%)	0868	0.6900	0.0060	0.7097	0.0367	0.0089	28	-0.54	132.00	0
132	Phenylalanine (%)	0859	0.6911	0.0025	0.7097	0.0367	0.0089	28	-0.51	132.00	0
132	Phenylalanine (%)	0910	0.6950	0.0100	0.7097	0.0367	0.0089	28	-0.40	132.00	0
132	Phenylalanine (%)	0652	0.7000	0.0200	0.7097	0.0367	0.0089	28	-0.27	132.00	0
132	Phenylalanine (%)	0148	0.7000	0.0200	0.7097	0.0367	0.0089	28	-0.27	132.05	0
132	Phenylalanine (%)	0889	0.7050	0.0000	0.7097	0.0367	0.0089	28	-0.13	132.99	0
132	Phenylalanine (%)	0878	0.7090	0.0040	0.7097	0.0367	0.0089	28	-0.02	132.00	0
132	Phenylalanine (%)	0676	0.7100	0.0400	0.7097	0.0367	0.0089	28	0.01	132.05	0
132	Phenylalanine (%)	0872	0.7145	0.0110	0.7097	0.0367	0.0089	28	0.13	132.00	0
132	Phenylalanine (%)	2059	0.7165	0.0010	0.7097	0.0367	0.0089	28	0.18	132.00	0
132	Phenylalanine (%)	0571	0.7180	0.0140	0.7097	0.0367	0.0089	28	0.22	132.00	0
132	Phenylalanine (%)	0941	0.7220	0.0140	0.7097	0.0367	0.0089	28	0.33	132.00	0
132	Phenylalanine (%)	0098	0.7225	0.0130	0.7097	0.0367	0.0089	28	0.35	132.02	0
132	Phenylalanine (%)	0227	0.7250	0.0100	0.7097	0.0367	0.0089	28	0.42	132.99	0
132	Phenylalanine (%)	0644	0.7305	0.0030	0.7097	0.0367	0.0089	28	0.57	132.00	0
132	Phenylalanine (%)	0354	0.7390	0.0040	0.7097	0.0367	0.0089	28	0.80	132.00	0
132	Phenylalanine (%)	0353	0.7450	0.0100	0.7097	0.0367	0.0089	28	0.96	132.00	0
132	Phenylalanine (%)	0870	0.7601	0.0009	0.7097	0.0367	0.0089	28	1.37	132.00	0
132	Phenylalanine (%)	0226	0.7704	0.0071	0.7097	0.0367	0.0089	28	1.65	132.00	0
132	Phenylalanine (%)	2196	0.7850	0.0000	0.7097	0.0367	0.0089	28	2.05	132.05	0
132	Phenylalanine (%)	0918	0.8045	0.0350	0.7097	0.0367	0.0089	28	2.58	132.00	0
132	Phenylalanine (%)	0407	0.7205	0.1150	0.7097	0.0367	0.0089	28	0.29	132.05	1
132	Phenylalanine (%)	2319	< 0.005		0.7097	0.0367	0.0089	28		132.05	5
133	Proline (%)	0859	0.1941	0.0036	0.9641	0.0680	0.0224	27	-11.32	133.00	0
133	Proline (%)	0675	0.7700	0.0200	0.9641	0.0680	0.0224	27	-2.85	133.00	0
133	Proline (%)	0889	0.7750	0.0400	0.9641	0.0680	0.0224	27	-2.78	133.99	0
133	Proline (%)	0910	0.9050	0.1100	0.9641	0.0680	0.0224	27	-0.87	133.00	0
133	Proline (%)	0870	0.9136	0.0040	0.9641	0.0680	0.0224	27	-0.74	133.00	0
133	Proline (%)	0354	0.9160	0.0560	0.9641	0.0680	0.0224	27	-0.71	133.00	0
133	Proline (%)	2059	0.9240	0.0100	0.9641	0.0680	0.0224	27	-0.59	133.00	0
133	Proline (%)	0652	0.9250	0.0100	0.9641	0.0680	0.0224	27	-0.58	133.00	0
133	Proline (%)	0676	0.9350	0.0100	0.9641	0.0680	0.0224	27	-0.43	133.05	0
133	Proline (%)	0619	0.9420	0.0280	0.9641	0.0680	0.0224	27	-0.33	133.00	0
133	Proline (%)	0868	0.9500	0.0160	0.9641	0.0680	0.0224	27	-0.21	133.00	0
133	Proline (%)	2146	0.9500	0.0200	0.9641	0.0680	0.0224	27	-0.21	133.99	0
133	Proline (%)	0878	0.9520	0.0100	0.9641	0.0680	0.0224	27	-0.18	133.00	0
133	Proline (%)	0941	0.9520	0.0040	0.9641	0.0680	0.0224	27	-0.18	133.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
133	Proline (%)	0571	0.9585	0.0390	0.9641	0.0680	0.0224	27	-0.08	133.00	0
133	Proline (%)	0968	0.9750	0.0100	0.9641	0.0680	0.0224	27	0.16	133.00	0
133	Proline (%)	0626	0.9780	0.0100	0.9641	0.0680	0.0224	27	0.20	133.05	0
133	Proline (%)	0227	0.9850	0.0100	0.9641	0.0680	0.0224	27	0.31	133.99	0
133	Proline (%)	0644	0.9920	0.0300	0.9641	0.0680	0.0224	27	0.41	133.00	0
133	Proline (%)	0148	0.9950	0.0100	0.9641	0.0680	0.0224	27	0.45	133.05	0
133	Proline (%)	0353	1.010	0.0000	0.9641	0.0680	0.0224	27	0.67	133.00	0
133	Proline (%)	0872	1.016	0.0150	0.9641	0.0680	0.0224	27	0.75	133.00	0
133	Proline (%)	0682	1.035	0.0000	0.9641	0.0680	0.0224	27	1.04	133.00	0
133	Proline (%)	0226	1.042	0.0022	0.9641	0.0680	0.0224	27	1.15	133.00	0
133	Proline (%)	0918	1.062	0.1361	0.9641	0.0680	0.0224	27	1.44	133.00	0
133	Proline (%)	2188	1.075	0.0000	0.9641	0.0680	0.0224	27	1.63	133.05	0
133	Proline (%)	2196	1.114	0.0000	0.9641	0.0680	0.0224	27	2.20	133.05	0
133	Proline (%)	0407	0.8015	0.4050	0.9641	0.0680	0.0224	27	-2.39	133.05	1
134	Serine (%)	0889	0.6200	0.0000	0.7197	0.0414	0.0128	28	-2.40	134.99	0
134	Serine (%)	0652	0.6300	0.0200	0.7197	0.0414	0.0128	28	-2.16	134.00	0
134	Serine (%)	0098	0.6465	0.0170	0.7197	0.0414	0.0128	28	-1.77	134.02	0
134	Serine (%)	0619	0.6800	0.0220	0.7197	0.0414	0.0128	28	-0.96	134.00	0
134	Serine (%)	0675	0.6900	0.0200	0.7197	0.0414	0.0128	28	-0.72	134.00	0
134	Serine (%)	2146	0.6900	0.0200	0.7197	0.0414	0.0128	28	-0.72	134.99	0
134	Serine (%)	0682	0.6910	0.0000	0.7197	0.0414	0.0128	28	-0.69	134.00	0
134	Serine (%)	0644	0.6945	0.0010	0.7197	0.0414	0.0128	28	-0.61	134.00	0
134	Serine (%)	0148	0.7050	0.0100	0.7197	0.0414	0.0128	28	-0.35	134.05	0
134	Serine (%)	0868	0.7055	0.0110	0.7197	0.0414	0.0128	28	-0.34	134.00	0
134	Serine (%)	0941	0.7085	0.0090	0.7197	0.0414	0.0128	28	-0.27	134.00	0
134	Serine (%)	0910	0.7100	0.0400	0.7197	0.0414	0.0128	28	-0.23	134.00	0
134	Serine (%)	0859	0.7142	0.0131	0.7197	0.0414	0.0128	28	-0.13	134.00	0
134	Serine (%)	0968	0.7150	0.0100	0.7197	0.0414	0.0128	28	-0.11	134.00	0
134	Serine (%)	0227	0.7200	0.0000	0.7197	0.0414	0.0128	28	0.01	134.99	0
134	Serine (%)	0878	0.7240	0.0020	0.7197	0.0414	0.0128	28	0.10	134.00	0
134	Serine (%)	0676	0.7250	0.0100	0.7197	0.0414	0.0128	28	0.13	134.05	0
134	Serine (%)	0571	0.7280	0.0060	0.7197	0.0414	0.0128	28	0.20	134.00	0
134	Serine (%)	0872	0.7315	0.0210	0.7197	0.0414	0.0128	28	0.29	134.00	0
134	Serine (%)	2059	0.7330	0.0120	0.7197	0.0414	0.0128	28	0.32	134.00	0
134	Serine (%)	2188	0.7380	0.0000	0.7197	0.0414	0.0128	28	0.44	134.05	0
134	Serine (%)	0918	0.7401	0.0105	0.7197	0.0414	0.0128	28	0.49	134.00	0
134	Serine (%)	0353	0.7500	0.0000	0.7197	0.0414	0.0128	28	0.73	134.00	0
134	Serine (%)	0354	0.7575	0.0050	0.7197	0.0414	0.0128	28	0.91	134.00	0
134	Serine (%)	0870	0.7870	0.0286	0.7197	0.0414	0.0128	28	1.62	134.00	0
134	Serine (%)	2196	0.7980	0.0000	0.7197	0.0414	0.0128	28	1.89	134.05	0
134	Serine (%)	0626	0.7985	0.0010	0.7197	0.0414	0.0128	28	1.90	134.05	0
134	Serine (%)	0226	0.8589	0.0685	0.7197	0.0414	0.0128	28	3.36	134.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
134	Serine (%)	0407	0.6920	0.0740	0.7197	0.0414	0.0128	28	-0.67	134.05	1
135	Threonine (%)	0889	0.4400	0.0000	0.5534	0.0207	0.0125	28	-5.48	135.99	0
135	Threonine (%)	0652	0.4600	0.0400	0.5534	0.0207	0.0125	28	-4.52	135.00	0
135	Threonine (%)	0407	0.4915	0.0270	0.5534	0.0207	0.0125	28	-2.99	135.05	0
135	Threonine (%)	0859	0.5267	0.0106	0.5534	0.0207	0.0125	28	-1.29	135.00	0
135	Threonine (%)	0619	0.5315	0.0090	0.5534	0.0207	0.0125	28	-1.06	135.00	0
135	Threonine (%)	0148	0.5350	0.0300	0.5534	0.0207	0.0125	28	-0.89	135.05	0
135	Threonine (%)	0682	0.5430	0.0000	0.5534	0.0207	0.0125	28	-0.50	135.00	0
135	Threonine (%)	0675	0.5450	0.0100	0.5534	0.0207	0.0125	28	-0.41	135.00	0
135	Threonine (%)	0868	0.5450	0.0100	0.5534	0.0207	0.0125	28	-0.41	135.00	0
135	Threonine (%)	0910	0.5450	0.0300	0.5534	0.0207	0.0125	28	-0.41	135.00	0
135	Threonine (%)	0968	0.5450	0.0100	0.5534	0.0207	0.0125	28	-0.41	135.00	0
135	Threonine (%)	0941	0.5470	0.0120	0.5534	0.0207	0.0125	28	-0.31	135.00	0
135	Threonine (%)	2146	0.5550	0.0100	0.5534	0.0207	0.0125	28	0.08	135.99	0
135	Threonine (%)	0872	0.5570	0.0020	0.5534	0.0207	0.0125	28	0.17	135.00	0
135	Threonine (%)	0571	0.5575	0.0110	0.5534	0.0207	0.0125	28	0.20	135.00	0
135	Threonine (%)	0878	0.5580	0.0100	0.5534	0.0207	0.0125	28	0.22	135.00	0
135	Threonine (%)	0644	0.5590	0.0000	0.5534	0.0207	0.0125	28	0.27	135.00	0
135	Threonine (%)	2188	0.5630	0.0000	0.5534	0.0207	0.0125	28	0.46	135.05	0
135	Threonine (%)	2059	0.5635	0.0070	0.5534	0.0207	0.0125	28	0.49	135.00	0
135	Threonine (%)	0626	0.5635	0.0070	0.5534	0.0207	0.0125	28	0.49	135.05	0
135	Threonine (%)	0918	0.5636	0.0446	0.5534	0.0207	0.0125	28	0.49	135.00	0
135	Threonine (%)	0098	0.5645	0.0010	0.5534	0.0207	0.0125	28	0.53	135.02	0
135	Threonine (%)	0353	0.5650	0.0100	0.5534	0.0207	0.0125	28	0.56	135.00	0
135	Threonine (%)	0227	0.5750	0.0100	0.5534	0.0207	0.0125	28	1.04	135.99	0
135	Threonine (%)	0870	0.5760	0.0162	0.5534	0.0207	0.0125	28	1.09	135.00	0
135	Threonine (%)	0354	0.5790	0.0100	0.5534	0.0207	0.0125	28	1.24	135.00	0
135	Threonine (%)	0226	0.5818	0.0213	0.5534	0.0207	0.0125	28	1.37	135.00	0
135	Threonine (%)	2196	0.5970	0.0000	0.5534	0.0207	0.0125	28	2.11	135.05	0
135	Threonine (%)	0676	0.5650	0.0700	0.5534	0.0207	0.0125	28	0.56	135.05	1
135	Threonine (%)	2319	< 0.005		0.5534	0.0207	0.0125	28		135.05	5
136	Tryptophan (%)	0407	0.1080	0.0120	0.1955	0.0234	0.0051	18	-3.75	136.05	0
136	Tryptophan (%)	0910	0.1650	0.0100	0.1955	0.0234	0.0051	18	-1.31	136.00	0
136	Tryptophan (%)	0941	0.1800	0.0100	0.1955	0.0234	0.0051	18	-0.66	136.00	0
136	Tryptophan (%)	2059	0.1825	0.0010	0.1955	0.0234	0.0051	18	-0.56	136.03	0
136	Tryptophan (%)	0878	0.1840	0.0020	0.1955	0.0234	0.0051	18	-0.49	136.01	0
136	Tryptophan (%)	0859	0.1846	0.0067	0.1955	0.0234	0.0051	18	-0.47	136.03	0
136	Tryptophan (%)	0868	0.1865	0.0110	0.1955	0.0234	0.0051	18	-0.39	136.01	0
136	Tryptophan (%)	0098	0.1905	0.0030	0.1955	0.0234	0.0051	18	-0.21	136.02	0
136	Tryptophan (%)	0872	0.1905	0.0030	0.1955	0.0234	0.0051	18	-0.21	136.03	0
136	Tryptophan (%)	0148	0.1930	0.0040	0.1955	0.0234	0.0051	18	-0.11	136.03	0
136	Tryptophan (%)	0644	0.1935	0.0010	0.1955	0.0234	0.0051	18	-0.09	136.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
136	Tryptophan (%)	0571	0.1935	0.0030	0.1955	0.0234	0.0051	18	-0.09	136.03	0
136	Tryptophan (%)	0619	0.1950	0.0040	0.1955	0.0234	0.0051	18	-0.02	136.00	0
136	Tryptophan (%)	0682	0.2170	0.0000	0.1955	0.0234	0.0051	18	0.92	136.00	0
136	Tryptophan (%)	0870	0.2170	0.0004	0.1955	0.0234	0.0051	18	0.92	136.00	0
136	Tryptophan (%)	0353	0.2250	0.0100	0.1955	0.0234	0.0051	18	1.26	136.03	0
136	Tryptophan (%)	0227	0.2350	0.0100	0.1955	0.0234	0.0051	18	1.69	136.00	0
136	Tryptophan (%)	0889	0.4300	0.0000	0.1955	0.0234	0.0051	18	10.04	136.99	0
136	Tryptophan (%)	0918	0.2050	0.0300	0.1955	0.0234	0.0051	18	0.41	136.00	1
137	Tyrosine (%)	0910	0.3550	0.0300	0.4860	0.0679	0.0136	22	-1.93	137.00	0
137	Tyrosine (%)	0682	0.3940	0.0000	0.4860	0.0679	0.0136	22	-1.36	137.00	0
137	Tyrosine (%)	0626	0.4045	0.0090	0.4860	0.0679	0.0136	22	-1.20	137.05	0
137	Tyrosine (%)	0889	0.4300	0.0000	0.4860	0.0679	0.0136	22	-0.83	137.99	0
137	Tyrosine (%)	0870	0.4324	0.0150	0.4860	0.0679	0.0136	22	-0.79	137.00	0
137	Tyrosine (%)	0354	0.4445	0.0170	0.4860	0.0679	0.0136	22	-0.61	137.00	0
137	Tyrosine (%)	0227	0.4450	0.0100	0.4860	0.0679	0.0136	22	-0.60	137.99	0
137	Tyrosine (%)	2146	0.4550	0.0100	0.4860	0.0679	0.0136	22	-0.46	137.99	0
137	Tyrosine (%)	0675	0.4600	0.0200	0.4860	0.0679	0.0136	22	-0.38	137.00	0
137	Tyrosine (%)	0968	0.4600	0.0000	0.4860	0.0679	0.0136	22	-0.38	137.00	0
137	Tyrosine (%)	0941	0.4775	0.0690	0.4860	0.0679	0.0136	22	-0.13	137.00	0
137	Tyrosine (%)	0098	0.4810	0.0220	0.4860	0.0679	0.0136	22	-0.07	137.02	0
137	Tyrosine (%)	0644	0.5010	0.0020	0.4860	0.0679	0.0136	22	0.22	137.00	0
137	Tyrosine (%)	0872	0.5010	0.0080	0.4860	0.0679	0.0136	22	0.22	137.00	0
137	Tyrosine (%)	2059	0.5355	0.0070	0.4860	0.0679	0.0136	22	0.73	137.00	0
137	Tyrosine (%)	2188	0.5410	0.0000	0.4860	0.0679	0.0136	22	0.81	137.05	0
137	Tyrosine (%)	0353	0.5450	0.0100	0.4860	0.0679	0.0136	22	0.87	137.00	0
137	Tyrosine (%)	0918	0.5466	0.0418	0.4860	0.0679	0.0136	22	0.89	137.00	0
137	Tyrosine (%)	0148	0.5550	0.0100	0.4860	0.0679	0.0136	22	1.02	137.05	0
137	Tyrosine (%)	0676	0.5550	0.0100	0.4860	0.0679	0.0136	22	1.02	137.05	0
137	Tyrosine (%)	0226	0.5568	0.0079	0.4860	0.0679	0.0136	22	1.04	137.00	0
137	Tyrosine (%)	2196	0.6380	0.0000	0.4860	0.0679	0.0136	22	2.24	137.05	0
137	Tyrosine (%)	0407	0.9385	0.5270	0.4860	0.0679	0.0136	22	6.66	137.05	1
137	Tyrosine (%)	2319	< 0.005		0.4860	0.0679	0.0136	22		137.05	5
138	Valine (%)	0226	0.5827	0.0636	0.7016	0.0379	0.0142	28	-3.14	138.00	0
138	Valine (%)	0353	0.6350	0.0100	0.7016	0.0379	0.0142	28	-1.76	138.00	0
138	Valine (%)	0889	0.6550	0.0000	0.7016	0.0379	0.0142	28	-1.23	138.99	0
138	Valine (%)	0626	0.6615	0.0030	0.7016	0.0379	0.0142	28	-1.06	138.05	0
138	Valine (%)	0354	0.6665	0.0610	0.7016	0.0379	0.0142	28	-0.93	138.00	0
138	Valine (%)	0859	0.6674	0.0140	0.7016	0.0379	0.0142	28	-0.90	138.00	0
138	Valine (%)	0652	0.6700	0.0400	0.7016	0.0379	0.0142	28	-0.83	138.00	0
138	Valine (%)	0868	0.6740	0.0080	0.7016	0.0379	0.0142	28	-0.73	138.00	0
138	Valine (%)	0682	0.6790	0.0000	0.7016	0.0379	0.0142	28	-0.60	138.00	0
138	Valine (%)	0870	0.6934	0.0150	0.7016	0.0379	0.0142	28	-0.22	138.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
138	Valine (%)	0619	0.6935	0.0010	0.7016	0.0379	0.0142	28	-0.21	138.00	0
138	Valine (%)	2146	0.6950	0.0100	0.7016	0.0379	0.0142	28	-0.17	138.99	0
138	Valine (%)	2188	0.6990	0.0000	0.7016	0.0379	0.0142	28	-0.07	138.05	0
138	Valine (%)	0941	0.7020	0.0080	0.7016	0.0379	0.0142	28	0.01	138.00	0
138	Valine (%)	0675	0.7050	0.0100	0.7016	0.0379	0.0142	28	0.09	138.00	0
138	Valine (%)	0872	0.7065	0.0210	0.7016	0.0379	0.0142	28	0.13	138.00	0
138	Valine (%)	0918	0.7066	0.0366	0.7016	0.0379	0.0142	28	0.13	138.00	0
138	Valine (%)	0878	0.7090	0.0020	0.7016	0.0379	0.0142	28	0.20	138.00	0
138	Valine (%)	0910	0.7150	0.0100	0.7016	0.0379	0.0142	28	0.35	138.00	0
138	Valine (%)	0968	0.7150	0.0100	0.7016	0.0379	0.0142	28	0.35	138.00	0
138	Valine (%)	0571	0.7235	0.0030	0.7016	0.0379	0.0142	28	0.58	138.00	0
138	Valine (%)	2059	0.7290	0.0060	0.7016	0.0379	0.0142	28	0.72	138.00	0
138	Valine (%)	0676	0.7400	0.0400	0.7016	0.0379	0.0142	28	1.01	138.05	0
138	Valine (%)	0098	0.7420	0.0100	0.7016	0.0379	0.0142	28	1.07	138.02	0
138	Valine (%)	0644	0.7450	0.0040	0.7016	0.0379	0.0142	28	1.15	138.00	0
138	Valine (%)	0148	0.7450	0.0100	0.7016	0.0379	0.0142	28	1.15	138.05	0
138	Valine (%)	2196	0.7590	0.0000	0.7016	0.0379	0.0142	28	1.52	138.05	0
138	Valine (%)	0227	0.7600	0.0000	0.7016	0.0379	0.0142	28	1.54	138.99	0
138	Valine (%)	0407	0.6360	0.1840	0.7016	0.0379	0.0142	28	-1.73	138.05	1
138	Valine (%)	2319	< 0.005		0.7016	0.0379	0.0142	28		138.05	5
139	Taurine (%)	0407	0.0055	0.0050				2		139.05	0
139	Taurine (%)	0682	0.0900	0.0000				2		139.00	0
139	Taurine (%)	0098	< 0.01					2		139.02	5
139	Taurine (%)	0227	< 0.01					2		139.99	5
150	Phytase (Units / kg)	0047	426.0	2.000	828.0	351.6	20.28	4		150.00	0
150	Phytase (Units / kg)	2022	676.5	17.00	828.0	351.6	20.28	4		150.99	0
150	Phytase (Units / kg)	0619	976.0	22.00	828.0	351.6	20.28	4		150.00	0
150	Phytase (Units / kg)	0218	1,233	40.10	828.0	351.6	20.28	4		150.00	0
150	Phytase (Units / kg)	0027	1,180	131.9	828.0	351.6	20.28	4		150.00	1
160	Fructose (%)	0619	0.1330	0.0020				3		160.99	0
160	Fructose (%)	0297	0.1385	0.0330				3		160.10	0
160	Fructose (%)	0227	0.2550	0.0100				3		160.99	0
161	Galactose (%)	0297	0.0000	0.0000				0		161.10	4
162	Glucose (%)	0619	0.0850	0.0000				3		162.99	0
162	Glucose (%)	0297	0.2000	0.0400				3		162.10	0
162	Glucose (%)	0227	0.3800	0.0000				3		162.99	0
162	Glucose (%)	0689	3.600	0.4000				3		162.99	2
163	Lactose (%)	0297	0.0000	0.0000				0		163.10	4
163	Lactose (%)	0619	0.0000	0.0000				0		163.99	4
163	Lactose (%)	0227	< 0.15					0		163.99	5
164	Maltose (%)	0297	0.2675	0.0370				1		164.10	0
164	Maltose (%)	0227	< 0.15					1		164.99	5

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
165	Sucrose (%)	0297	1.783	0.1130				3		165.10	0
165	Sucrose (%)	0619	1.840	0.0400				3		165.99	0
165	Sucrose (%)	0227	2.080	0.0200				3		165.99	0
166	Raffinose (%)	0619	0.3395	0.0030				3		166.99	0
166	Raffinose (%)	0297	0.3610	0.0280				3		166.10	0
166	Raffinose (%)	0227	0.3900	0.0000				3		166.99	0
167	Stachyose (%)	0619	0.6895	0.0050				3		167.99	0
167	Stachyose (%)	0227	0.7050	0.0100				3		167.99	0
167	Stachyose (%)	0297	0.7150	0.0320				3		167.10	0
351	Chlortetracycline (ppm)	2314	0.8604	0.0046	2.590	1.301	0.2486	6	-1.33	351.05	0
351	Chlortetracycline (ppm)	2055	1.800	0.2000	2.590	1.301	0.2486	6	-0.61	351.05	0
351	Chlortetracycline (ppm)	0148	2.665	0.9300	2.590	1.301	0.2486	6	0.06	351.05	0
351	Chlortetracycline (ppm)	0218	2.832	0.0270	2.590	1.301	0.2486	6	0.19	351.00	0
351	Chlortetracycline (ppm)	0939	3.205	0.2700	2.590	1.301	0.2486	6	0.47	351.05	0
351	Chlortetracycline (ppm)	0619	4.180	0.0600	2.590	1.301	0.2486	6	1.22	351.05	0
382	Sulfamethazine (ppm)	2234	2.090	0.1600				3		382.04	0
382	Sulfamethazine (ppm)	2314	2.947	0.1115				3		382.04	0
382	Sulfamethazine (ppm)	0148	4.440	0.6600				3		382.04	0
386	Tiamulin (ppm)	2144	62.92	0.3500	114.4	17.59	4.764	8	-2.92	386.02	0
386	Tiamulin (ppm)	0668	105.5	3.000	114.4	17.59	4.764	8	-0.50	386.00	0
386	Tiamulin (ppm)	2192	109.5	2.900	114.4	17.59	4.764	8	-0.28	386.00	0
386	Tiamulin (ppm)	0148	111.1	1.822	114.4	17.59	4.764	8	-0.19	386.02	0
386	Tiamulin (ppm)	2234	112.4	7.440	114.4	17.59	4.764	8	-0.11	386.02	0
386	Tiamulin (ppm)	0910	122.2	8.600	114.4	17.59	4.764	8	0.45	386.00	0
386	Tiamulin (ppm)	0001	125.5	13.00	114.4	17.59	4.764	8	0.63	386.00	0
386	Tiamulin (ppm)	0619	153.5	1.000	114.4	17.59	4.764	8	2.23	386.02	0
388	Tylosin (ppm)	2146	44.84	0.7500	58.22	13.91	0.7189	9	-0.96	388.99	0
388	Tylosin (ppm)	0875	45.00	0.0000	58.22	13.91	0.7189	9	-0.95	388.05	0
388	Tylosin (ppm)	2144	49.90	0.1400	58.22	13.91	0.7189	9	-0.60	388.05	0
388	Tylosin (ppm)	0910	52.50	1.000	58.22	13.91	0.7189	9	-0.41	388.03	0
388	Tylosin (ppm)	2192	53.50	0.9800	58.22	13.91	0.7189	9	-0.34	388.03	0
388	Tylosin (ppm)	0148	62.40	0.8000	58.22	13.91	0.7189	9	0.30	388.05	0
388	Tylosin (ppm)	0032	62.65	1.300	58.22	13.91	0.7189	9	0.32	388.03	0
388	Tylosin (ppm)	2234	74.13	0.8000	58.22	13.91	0.7189	9	1.14	388.05	0
388	Tylosin (ppm)	0619	89.75	0.7000	58.22	13.91	0.7189	9	2.27	388.05	0
388	Tylosin (ppm)	0033	97.80	14.40	58.22	13.91	0.7189	9	2.84	388.03	1
400	Water Activity (Units)	2174	0.5200	0.0200	0.5402	0.0148	0.0055	11	-1.36	400.01	0
400	Water Activity (Units)	2168	0.5285	0.0030	0.5402	0.0148	0.0055	11	-0.79	400.01	0
400	Water Activity (Units)	0407	0.5311	0.0076	0.5402	0.0148	0.0055	11	-0.61	400.01	0
400	Water Activity (Units)	0942	0.5355	0.0010	0.5402	0.0148	0.0055	11	-0.32	400.01	0
400	Water Activity (Units)	0843	0.5355	0.0010	0.5402	0.0148	0.0055	11	-0.32	400.99	0
400	Water Activity (Units)	2073	0.5355	0.0070	0.5402	0.0148	0.0055	11	-0.32	400.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
400	Water Activity (Units)	0843	0.5410	0.0020	0.5402	0.0148	0.0055	11	0.06	400.01	0
400	Water Activity (Units)	2268	0.5430	0.0000	0.5402	0.0148	0.0055	11	0.19	400.01	0
400	Water Activity (Units)	0589	0.5480	0.0060	0.5402	0.0148	0.0055	11	0.53	400.01	0
400	Water Activity (Units)	2303	0.5613	0.0084	0.5402	0.0148	0.0055	11	1.43	400.01	0
400	Water Activity (Units)	2109	0.5972	0.0050	0.5402	0.0148	0.0055	11	3.86	400.01	0
516	Arsenic, Total (ppm)	0045	0.1320	0.0020	0.1471	0.0155	0.0037	10	-0.97	516.00	0
516	Arsenic, Total (ppm)	0186	0.1335	0.0030	0.1471	0.0155	0.0037	10	-0.88	516.52	0
516	Arsenic, Total (ppm)	0553	0.1360	0.0060	0.1471	0.0155	0.0037	10	-0.72	516.53	0
516	Arsenic, Total (ppm)	0171	0.1400	0.0040	0.1471	0.0155	0.0037	10	-0.46	516.00	0
516	Arsenic, Total (ppm)	0227	0.1440	0.0120	0.1471	0.0155	0.0037	10	-0.20	516.53	0
516	Arsenic, Total (ppm)	0910	0.1450	0.0100	0.1471	0.0155	0.0037	10	-0.13	516.52	0
516	Arsenic, Total (ppm)	0939	0.1500	0.0000	0.1471	0.0155	0.0037	10	0.19	516.53	0
516	Arsenic, Total (ppm)	2033	0.1500	0.0000	0.1471	0.0155	0.0037	10	0.19	516.53	0
516	Arsenic, Total (ppm)	2053	0.1700	0.0000	0.1471	0.0155	0.0037	10	1.48	516.52	0
516	Arsenic, Total (ppm)	0918	0.2038	0.0000	0.1471	0.0155	0.0037	10	3.66	516.53	0
516	Arsenic, Total (ppm)	0407	1.057	0.1200	0.1471	0.0155	0.0037	10	58.70	516.43	1
516	Arsenic, Total (ppm)	2302	0.0000	0.0000	0.1471	0.0155	0.0037	10		516.99	4
516	Arsenic, Total (ppm)	0682	< 10		0.1471	0.0155	0.0037	10		516.43	5
518	Cadmium (ppm)	0171	0.0960	0.0000	0.1123	0.0155	0.0055	12	-1.05	518.41	0
518	Cadmium (ppm)	0227	0.0975	0.0030	0.1123	0.0155	0.0055	12	-0.96	518.53	0
518	Cadmium (ppm)	2033	0.1000	0.0200	0.1123	0.0155	0.0055	12	-0.79	518.53	0
518	Cadmium (ppm)	0939	0.1050	0.0100	0.1123	0.0155	0.0055	12	-0.47	518.53	0
518	Cadmium (ppm)	0186	0.1060	0.0040	0.1123	0.0155	0.0055	12	-0.41	518.52	0
518	Cadmium (ppm)	0553	0.1085	0.0030	0.1123	0.0155	0.0055	12	-0.25	518.53	0
518	Cadmium (ppm)	2053	0.1100	0.0000	0.1123	0.0155	0.0055	12	-0.15	518.52	0
518	Cadmium (ppm)	0918	0.1115	0.0050	0.1123	0.0155	0.0055	12	-0.05	518.53	0
518	Cadmium (ppm)	0910	0.1150	0.0100	0.1123	0.0155	0.0055	12	0.17	518.52	0
518	Cadmium (ppm)	0968	0.1270	0.0040	0.1123	0.0155	0.0055	12	0.95	518.43	0
518	Cadmium (ppm)	0407	0.1437	0.0067	0.1123	0.0155	0.0055	12	2.02	518.43	0
518	Cadmium (ppm)	2246	0.2300	0.0000	0.1123	0.0155	0.0055	12	7.60	518.34	0
518	Cadmium (ppm)	0407	0.1134	0.0318	0.1123	0.0155	0.0055	12	0.07	518.41	1
518	Cadmium (ppm)	2302	0.0000	0.0000	0.1123	0.0155	0.0055	12		518.99	4
518	Cadmium (ppm)	0689	< 0.2		0.1123	0.0155	0.0055	12		518.31	5
518	Cadmium (ppm)	0682	< 0.2		0.1123	0.0155	0.0055	12		518.43	5
520	Chromium (ppm)	0171	1.893	0.0000	3.419	1.283	0.0804	13	-1.19	520.41	0
520	Chromium (ppm)	0186	2.235	0.0300	3.419	1.283	0.0804	13	-0.92	520.52	0
520	Chromium (ppm)	0407	2.265	0.2420	3.419	1.283	0.0804	13	-0.90	520.41	0
520	Chromium (ppm)	0918	2.326	0.0574	3.419	1.283	0.0804	13	-0.85	520.53	0
520	Chromium (ppm)	0968	2.451	0.0640	3.419	1.283	0.0804	13	-0.75	520.43	0
520	Chromium (ppm)	0407	2.724	0.0763	3.419	1.283	0.0804	13	-0.54	520.43	0
520	Chromium (ppm)	0553	3.995	0.1500	3.419	1.283	0.0804	13	0.45	520.53	0
520	Chromium (ppm)	0164	4.035	0.0500	3.419	1.283	0.0804	13	0.48	520.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
520	Chromium (ppm)	0693	4.054	0.1460	3.419	1.283	0.0804	13	0.50	520.42	0
520	Chromium (ppm)	0045	4.215	0.0500	3.419	1.283	0.0804	13	0.62	520.42	0
520	Chromium (ppm)	0939	4.250	0.1200	3.419	1.283	0.0804	13	0.65	520.53	0
520	Chromium (ppm)	0682	4.660	0.0000	3.419	1.283	0.0804	13	0.97	520.43	0
520	Chromium (ppm)	0510	5.490	0.0600	3.419	1.283	0.0804	13	1.61	520.43	0
526	Lead (ppm)	0171	0.0880	0.0100	0.1949	0.0249	0.0138	11	-4.29	526.41	0
526	Lead (ppm)	0968	0.1800	0.0100	0.1949	0.0249	0.0138	11	-0.60	526.43	0
526	Lead (ppm)	2053	0.1800	0.0000	0.1949	0.0249	0.0138	11	-0.60	526.52	0
526	Lead (ppm)	0227	0.1800	0.0140	0.1949	0.0249	0.0138	11	-0.60	526.53	0
526	Lead (ppm)	0553	0.1915	0.0150	0.1949	0.0249	0.0138	11	-0.14	526.53	0
526	Lead (ppm)	0918	0.1919	0.0013	0.1949	0.0249	0.0138	11	-0.12	526.53	0
526	Lead (ppm)	0407	0.1959	0.0125	0.1949	0.0249	0.0138	11	0.04	526.41	0
526	Lead (ppm)	0910	0.2000	0.0000	0.1949	0.0249	0.0138	11	0.20	526.52	0
526	Lead (ppm)	2033	0.2050	0.0300	0.1949	0.0249	0.0138	11	0.40	526.53	0
526	Lead (ppm)	0939	0.2300	0.0400	0.1949	0.0249	0.0138	11	1.41	526.53	0
526	Lead (ppm)	0407	0.2563	0.0187	0.1949	0.0249	0.0138	11	2.46	526.43	0
526	Lead (ppm)	2246	0.4750	0.1300	0.1949	0.0249	0.0138	11	11.23	526.34	1
526	Lead (ppm)	2302	0.0000	0.0000	0.1949	0.0249	0.0138	11		526.99	4
526	Lead (ppm)	0689	< 0.2		0.1949	0.0249	0.0138	11		526.31	5
526	Lead (ppm)	0682	< 3		0.1949	0.0249	0.0138	11		526.43	5
529	Mercury (ppb)	2302	0.0000	0.0000				0		529.99	4
529	Mercury (ppb)	0227	< 10					0		529.99	5
529	Mercury (ppb)	0682	< 10000					0		529.99	5
539	Nickel (ppm)	0171	1.860	0.0200	2.434	0.5133	0.0456	8	-1.12	539.41	0
539	Nickel (ppm)	0407	2.021	0.0247	2.434	0.5133	0.0456	8	-0.80	539.41	0
539	Nickel (ppm)	0186	2.065	0.0300	2.434	0.5133	0.0456	8	-0.72	539.52	0
539	Nickel (ppm)	0407	2.165	0.1276	2.434	0.5133	0.0456	8	-0.52	539.43	0
539	Nickel (ppm)	0553	2.635	0.1100	2.434	0.5133	0.0456	8	0.39	539.53	0
539	Nickel (ppm)	2033	2.845	0.0500	2.434	0.5133	0.0456	8	0.80	539.43	0
539	Nickel (ppm)	0682	2.930	0.0000	2.434	0.5133	0.0456	8	0.97	539.43	0
539	Nickel (ppm)	0918	2.954	0.0027	2.434	0.5133	0.0456	8	1.01	539.53	0
702	Butyric Acid (4:0) (%)	0619	0.0000	0.0000				0		702.00	4
704	Caproic Acid (6:0) (%)	0619	0.0115	0.0010				1		704.00	0
706	Caprylic acid (8:0) (%) (w/w)	0619	0.0105	0.0010				1		706.01	0
708	Capric acid (10:0) (%) (w/w)	0619	0.0155	0.0010				1		708.01	0
710	Lauric Acid (12:0) (%) (w/w)	0676	0.0015	0.0010				3		710.99	0
710	Lauric Acid (12:0) (%) (w/w)	0164	0.0125	0.0010				3		710.99	0
710	Lauric Acid (12:0) (%) (w/w)	0619	0.0230	0.0020				3		710.01	0
710	Lauric Acid (12:0) (%) (w/w)	0226	< 0.005					3		710.99	5
714	Myristic Acid (14:0) (%) (w/w)	0226	0.0050	0.0000				3		714.99	0
714	Myristic Acid (14:0) (%) (w/w)	0164	0.0160	0.0000				3		714.99	0
714	Myristic Acid (14:0) (%) (w/w)	0619	0.0305	0.0030				3		714.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
716	Palmitic Acid (16:0) (% (w/w))	0619	0.5075	0.0070				3	716.01	0	
716	Palmitic Acid (16:0) (% (w/w))	0226	0.5200	0.0060				3	716.99	0	
716	Palmitic Acid (16:0) (% (w/w))	0164	0.7220	0.0220				3	716.99	0	
718	Palmitoleic Acid (9c-16:1) (% (w/w))	0226	0.0080	0.0000				3	718.99	0	
718	Palmitoleic Acid (9c-16:1) (% (w/w))	0676	0.0090	0.0000				3	718.99	0	
718	Palmitoleic Acid (9c-16:1) (% (w/w))	0619	0.0415	0.0050				3	718.01	0	
722	Stearic Acid (18:0) (% (w/w))	0619	0.0805	0.0030				3	722.01	0	
722	Stearic Acid (18:0) (% (w/w))	0226	0.0885	0.0010				3	722.99	0	
722	Stearic Acid (18:0) (% (w/w))	0164	0.1170	0.0020				3	722.99	0	
724	Oleic Acid (9c-18:1) (% (w/w))	0619	0.6170	0.0000				3	724.01	0	
724	Oleic Acid (9c-18:1) (% (w/w))	0226	0.7025	0.0050				3	724.99	0	
724	Oleic Acid (9c-18:1) (% (w/w))	0164	0.8340	0.0220				3	724.99	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0148	1.345	0.0100	1.446	0.0970	0.0216	5	726.02	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0676	1.380	0.0090	1.446	0.0970	0.0216	5	726.99	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0226	1.432	0.0370	1.446	0.0970	0.0216	5	726.99	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0619	1.480	0.0000	1.446	0.0970	0.0216	5	726.01	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0164	1.593	0.0520	1.446	0.0970	0.0216	5	726.99	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v)	0676	0.0630	0.0000	0.0699	0.0066	0.0018	4	728.99	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v)	0226	0.0665	0.0030	0.0699	0.0066	0.0018	4	728.99	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v)	0619	0.0720	0.0040	0.0699	0.0066	0.0018	4	728.01	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v)	0164	0.0780	0.0000	0.0699	0.0066	0.0018	4	728.99	0	
730	Arachidic Acid (20:0) (% (w/w))	0226	0.0110	0.0000				3	730.99	0	
730	Arachidic Acid (20:0) (% (w/w))	0164	0.0140	0.0000				3	730.99	0	
730	Arachidic Acid (20:0) (% (w/w))	0619	0.0410	0.0040				3	730.01	0	
732	Gondoic Acid (11c-20:1) (% (w/w))	0226	0.0115	0.0010				3	732.99	0	
732	Gondoic Acid (11c-20:1) (% (w/w))	0164	0.0130	0.0000				3	732.99	0	
732	Gondoic Acid (11c-20:1) (% (w/w))	0619	0.0440	0.0040				3	732.01	0	
736	Arachidonic Acid (5c,8c,11c,14c-20:4) (% (w/v)	0676	0.0020	0.0000				1	736.99	0	
736	Arachidonic Acid (5c,8c,11c,14c-20:4) (% (w/v)	0619	0.0000	0.0000				1	736.01	4	
738	Mead Acid (11c,14c,17c-20:3) (% (w/w))	0619	0.0000	0.0000				0	738.01	4	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c,19c-20:5) (% (w/w))	0619	0.0000	0.0000				0	740.01	4	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c,19c-20:5) (% (w/w))	0226	0.0000	0.0000				0	740.99	4	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c,19c-20:5) (% (w/w))	0676	0.0000	0.0000				0	740.99	4	
742	Behenic Acid (22:0) (% (w/w))	0226	0.0065	0.0010				2	742.99	0	
742	Behenic Acid (22:0) (% (w/w))	0164	0.0130	0.0000				2	742.99	0	
744	Erucic Acid (13c-22:1) (% (w/w))	0619	0.0000	0.0000				0	744.01	4	
744	Erucic Acid (13c-22:1) (% (w/w))	0226	< 0.005					0	744.99	5	
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c-22:5) (% (w/w))	0619	0.0000	0.0000				0	746.01	4	
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c-22:5) (% (w/w))	0226	0.0000	0.0000				0	746.99	4	
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c-22:5) (% (w/w))	0676	0.0000	0.0000				0	746.99	4	
748	Lignoceric Acid (24:0) (% (w/w))	0226	0.0090	0.0000				1	748.99	0	
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22c-24:6) (% (w/w))	0619	0.0000	0.0000				0	750.01	4	

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16	0226	0.0000	0.0000				0	750.99	4	
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16	0676	0.0000	0.0000				0	750.99	4	
752	Nervonic Acid (24:1) isomers (% (w/w))	0619	0.0000	0.0000				0	752.01	4	
752	Nervonic Acid (24:1) isomers (% (w/w))	0226	< 0.005					0	752.99	5	
754	Total n-3 Polyunsaturated (Omega-3) Fatty Aci	0676	0.0600	0.0000				3	754.99	0	
754	Total n-3 Polyunsaturated (Omega-3) Fatty Aci	0148	0.0748	0.0016				3	754.02	0	
754	Total n-3 Polyunsaturated (Omega-3) Fatty Aci	0164	0.0835	0.0010				3	754.99	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Aci	0148	1.345	0.0100				3	756.01	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Aci	0676	1.385	0.0100				3	756.99	0	
756	Total n-6 Polyunsaturated (Omega-6) Fatty Aci	0164	1.605	0.0540				3	756.99	0	
758	Total Saturated Fatty Acids (% (w/w))	0164	0.9225	0.0310				1	758.99	0	
762	Total Monounsaturated Fatty Acids (% (w/w))	0164	0.9150	0.0240				1	762.99	0	
766	Total Polyunsaturated Fatty Acids (% (w/w))	0164	1.696	0.0520				1	766.99	0	
770	Total Fat (equivalent to NLEA) (% (w/w))	0164	3.735	0.1140				1	770.99	0	
772	Total Fatty Acids (% (w/w))	0226	2.860	0.0540				2	772.99	0	
772	Total Fatty Acids (% (w/w))	0164	3.571	0.1090				2	772.99	0	

Notes: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2 and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = Rejected for duplicates too far apart, 2 = Rejected as outlier, 5 = A reporting limit and 4 = zeros submitted as values. Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). Flag 3 indicates not used in statistics.