



**AAFCO**  
Proficiency Testing Program



**Mycotoxin Contaminants Scheme**

**Puppy Feed**

**Test Material Code # 202063**

**# Labs Reporting: 59**

**# Analytes Reported 15**

**Issue Date : 08/31/2020**

**Analyte Proficiency Testing Report**

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
600	Total Aflatoxin (ppb)	0035	7.154	0.1310	20.63	6.999	1.490	43	-1.93	600.13	0
600	Total Aflatoxin (ppb)	0619	12.10	1.800	20.63	6.999	1.490	43	-1.22	600.24	0
600	Total Aflatoxin (ppb)	2113	12.75	3.100	20.63	6.999	1.490	43	-1.13	600.01	0
600	Total Aflatoxin (ppb)	0202	13.05	0.5000	20.63	6.999	1.490	43	-1.08	600.12	0
600	Total Aflatoxin (ppb)	0511	13.70	0.0000	20.63	6.999	1.490	43	-0.99	600.01	0
600	Total Aflatoxin (ppb)	2033	14.30	2.000	20.63	6.999	1.490	43	-0.90	600.24	0
600	Total Aflatoxin (ppb)	2066	15.55	0.0300	20.63	6.999	1.490	43	-0.73	600.12	0
600	Total Aflatoxin (ppb)	0027	15.65	1.500	20.63	6.999	1.490	43	-0.71	600.01	0
600	Total Aflatoxin (ppb)	0034	15.75	0.1000	20.63	6.999	1.490	43	-0.70	600.01	0
600	Total Aflatoxin (ppb)	0529	16.10	0.8000	20.63	6.999	1.490	43	-0.65	600.01	0
600	Total Aflatoxin (ppb)	2050	16.30	1.000	20.63	6.999	1.490	43	-0.62	600.98	0
600	Total Aflatoxin (ppb)	0004	17.50	1.000	20.63	6.999	1.490	43	-0.45	600.14	0
600	Total Aflatoxin (ppb)	0010	17.50	1.000	20.63	6.999	1.490	43	-0.45	600.20	0
600	Total Aflatoxin (ppb)	2192	18.85	2.270	20.63	6.999	1.490	43	-0.26	600.20	0
600	Total Aflatoxin (ppb)	0142	18.90	2.400	20.63	6.999	1.490	43	-0.25	600.98	0
600	Total Aflatoxin (ppb)	2066	19.45	0.3000	20.63	6.999	1.490	43	-0.17	600.20	0
600	Total Aflatoxin (ppb)	0563	19.50	1.000	20.63	6.999	1.490	43	-0.16	600.20	0
600	Total Aflatoxin (ppb)	0674	19.70	0.0000	20.63	6.999	1.490	43	-0.13	600.20	0
600	Total Aflatoxin (ppb)	0407	19.77	0.8200	20.63	6.999	1.490	43	-0.12	600.13	0
600	Total Aflatoxin (ppb)	0682	20.00	0.0000	20.63	6.999	1.490	43	-0.09	600.22	0
600	Total Aflatoxin (ppb)	2129	20.00	1.200	20.63	6.999	1.490	43	-0.09	600.24	0
600	Total Aflatoxin (ppb)	2096	20.25	4.300	20.63	6.999	1.490	43	-0.05	600.01	0
600	Total Aflatoxin (ppb)	0554	20.40	0.8000	20.63	6.999	1.490	43	-0.03	600.01	0
600	Total Aflatoxin (ppb)	0190	20.50	3.000	20.63	6.999	1.490	43	-0.02	600.14	0
600	Total Aflatoxin (ppb)	0656	22.22	0.0000	20.63	6.999	1.490	43	0.23	600.12	0
600	Total Aflatoxin (ppb)	0590	22.25	7.700	20.63	6.999	1.490	43	0.23	600.12	0
600	Total Aflatoxin (ppb)	2226	22.27	1.472	20.63	6.999	1.490	43	0.23	600.12	0
600	Total Aflatoxin (ppb)	0066	22.75	0.1000	20.63	6.999	1.490	43	0.30	600.13	0
600	Total Aflatoxin (ppb)	0208	22.80	0.0000	20.63	6.999	1.490	43	0.31	600.01	0
600	Total Aflatoxin (ppb)	2268	23.50	5.000	20.63	6.999	1.490	43	0.41	600.14	0
600	Total Aflatoxin (ppb)	0876	23.60	0.2000	20.63	6.999	1.490	43	0.42	600.01	0
600	Total Aflatoxin (ppb)	0629	23.75	0.0400	20.63	6.999	1.490	43	0.45	600.20	0
600	Total Aflatoxin (ppb)	0051	24.00	0.2000	20.63	6.999	1.490	43	0.48	600.20	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
600	Total Aflatoxin (ppb)	2269	24.47	0.1900	20.63	6.999	1.490	43	0.55	600.98	0
600	Total Aflatoxin (ppb)	2098	24.80	1.800	20.63	6.999	1.490	43	0.60	600.01	0
600	Total Aflatoxin (ppb)	0870	25.50	2.000	20.63	6.999	1.490	43	0.70	600.98	0
600	Total Aflatoxin (ppb)	2244	26.26	3.692	20.63	6.999	1.490	43	0.80	600.20	0
600	Total Aflatoxin (ppb)	2244	26.32	0.1080	20.63	6.999	1.490	43	0.81	600.98	0
600	Total Aflatoxin (ppb)	0032	27.05	3.300	20.63	6.999	1.490	43	0.92	600.24	0
600	Total Aflatoxin (ppb)	0160	29.35	3.300	20.63	6.999	1.490	43	1.25	600.24	0
600	Total Aflatoxin (ppb)	2057	29.36	1.235	20.63	6.999	1.490	43	1.25	600.24	0
600	Total Aflatoxin (ppb)	0098	30.07	2.690	20.63	6.999	1.490	43	1.35	600.22	0
600	Total Aflatoxin (ppb)	0353	39.80	2.000	20.63	6.999	1.490	43	2.74	600.24	0
601	Aflatoxin B1 (ppb)	0619	11.70	2.600	22.97	7.768	1.442	31	-1.45	601.23	0
601	Aflatoxin B1 (ppb)	2033	13.15	1.700	22.97	7.768	1.442	31	-1.26	601.23	0
601	Aflatoxin B1 (ppb)	2192	17.60	2.150	22.97	7.768	1.442	31	-0.69	601.20	0
601	Aflatoxin B1 (ppb)	2053	17.85	2.300	22.97	7.768	1.442	31	-0.66	601.24	0
601	Aflatoxin B1 (ppb)	2066	18.15	0.3000	22.97	7.768	1.442	31	-0.62	601.20	0
601	Aflatoxin B1 (ppb)	0682	18.50	0.0000	22.97	7.768	1.442	31	-0.57	601.24	0
601	Aflatoxin B1 (ppb)	0009	19.36	0.1010	22.97	7.768	1.442	31	-0.46	601.23	0
601	Aflatoxin B1 (ppb)	0563	19.50	1.000	22.97	7.768	1.442	31	-0.45	601.20	0
601	Aflatoxin B1 (ppb)	2057	21.91	0.4466	22.97	7.768	1.442	31	-0.14	601.23	0
601	Aflatoxin B1 (ppb)	0868	22.10	0.2000	22.97	7.768	1.442	31	-0.11	601.21	0
601	Aflatoxin B1 (ppb)	2239	22.15	1.300	22.97	7.768	1.442	31	-0.11	601.98	0
601	Aflatoxin B1 (ppb)	0629	22.21	0.0300	22.97	7.768	1.442	31	-0.10	601.20	0
601	Aflatoxin B1 (ppb)	0553	22.39	3.540	22.97	7.768	1.442	31	-0.07	601.23	0
601	Aflatoxin B1 (ppb)	2207	22.50	1.000	22.97	7.768	1.442	31	-0.06	601.23	0
601	Aflatoxin B1 (ppb)	0910	22.74	1.270	22.97	7.768	1.442	31	-0.03	601.21	0
601	Aflatoxin B1 (ppb)	2244	22.78	0.7800	22.97	7.768	1.442	31	-0.02	601.98	0
601	Aflatoxin B1 (ppb)	0904	22.99	1.690	22.97	7.768	1.442	31	0.00	601.23	0
601	Aflatoxin B1 (ppb)	2244	23.17	3.872	22.97	7.768	1.442	31	0.03	601.20	0
601	Aflatoxin B1 (ppb)	0870	23.60	3.520	22.97	7.768	1.442	31	0.08	601.23	0
601	Aflatoxin B1 (ppb)	0964	24.00	0.0000	22.97	7.768	1.442	31	0.13	601.23	0
601	Aflatoxin B1 (ppb)	2269	24.53	0.3800	22.97	7.768	1.442	31	0.20	601.98	0
601	Aflatoxin B1 (ppb)	0878	24.55	0.3000	22.97	7.768	1.442	31	0.20	601.98	0
601	Aflatoxin B1 (ppb)	0043	25.43	0.2500	22.97	7.768	1.442	31	0.32	601.23	0
601	Aflatoxin B1 (ppb)	2055	25.70	0.6000	22.97	7.768	1.442	31	0.35	601.23	0
601	Aflatoxin B1 (ppb)	2140	26.00	2.000	22.97	7.768	1.442	31	0.39	601.23	0
601	Aflatoxin B1 (ppb)	0032	27.05	3.300	22.97	7.768	1.442	31	0.53	601.23	0
601	Aflatoxin B1 (ppb)	0868	27.65	0.3000	22.97	7.768	1.442	31	0.60	601.98	0
601	Aflatoxin B1 (ppb)	0160	27.67	3.480	22.97	7.768	1.442	31	0.61	601.23	0
601	Aflatoxin B1 (ppb)	0098	28.05	2.470	22.97	7.768	1.442	31	0.65	601.24	0
601	Aflatoxin B1 (ppb)	0353	37.00	2.000	22.97	7.768	1.442	31	1.81	601.23	0
601	Aflatoxin B1 (ppb)	2155	47.94	1.820	22.97	7.768	1.442	31	3.21	601.23	0
602	Aflatoxin B2 (ppb)	2033	1.150	0.3000	1.738	0.6306	0.1253	18	-0.93	602.23	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
602	Aflatoxin B2 (ppb)	0868	1.245	0.0300	1.738	0.6306	0.1253	18	-0.78	602.21	0
602	Aflatoxin B2 (ppb)	2053	1.250	0.1000	1.738	0.6306	0.1253	18	-0.77	602.24	0
602	Aflatoxin B2 (ppb)	0043	1.290	0.0460	1.738	0.6306	0.1253	18	-0.71	602.23	0
602	Aflatoxin B2 (ppb)	0682	1.300	0.0000	1.738	0.6306	0.1253	18	-0.70	602.24	0
602	Aflatoxin B2 (ppb)	2066	1.310	0.1200	1.738	0.6306	0.1253	18	-0.68	602.20	0
602	Aflatoxin B2 (ppb)	2192	1.360	0.1400	1.738	0.6306	0.1253	18	-0.60	602.20	0
602	Aflatoxin B2 (ppb)	0629	1.545	0.0700	1.738	0.6306	0.1253	18	-0.31	602.20	0
602	Aflatoxin B2 (ppb)	0910	1.660	0.1000	1.738	0.6306	0.1253	18	-0.12	602.21	0
602	Aflatoxin B2 (ppb)	0160	1.690	0.2000	1.738	0.6306	0.1253	18	-0.08	602.23	0
602	Aflatoxin B2 (ppb)	0870	1.760	0.1200	1.738	0.6306	0.1253	18	0.03	602.23	0
602	Aflatoxin B2 (ppb)	0904	1.815	0.0900	1.738	0.6306	0.1253	18	0.12	602.23	0
602	Aflatoxin B2 (ppb)	0098	2.020	0.2200	1.738	0.6306	0.1253	18	0.45	602.24	0
602	Aflatoxin B2 (ppb)	0964	2.075	0.2100	1.738	0.6306	0.1253	18	0.53	602.23	0
602	Aflatoxin B2 (ppb)	0553	2.135	0.0100	1.738	0.6306	0.1253	18	0.63	602.23	0
602	Aflatoxin B2 (ppb)	2244	2.642	0.0510	1.738	0.6306	0.1253	18	1.43	602.20	0
602	Aflatoxin B2 (ppb)	0353	2.800	0.2000	1.738	0.6306	0.1253	18	1.68	602.23	0
602	Aflatoxin B2 (ppb)	2057	3.170	0.2487	1.738	0.6306	0.1253	18	2.27	602.23	0
602	Aflatoxin B2 (ppb)	0032	0.0000	0.0000	1.738	0.6306	0.1253	18		602.23	4
602	Aflatoxin B2 (ppb)	0619	0.3680	0.7360	1.738	0.6306	0.1253	18		602.23	4
602	Aflatoxin B2 (ppb)	2207	< 1		1.738	0.6306	0.1253	18		602.23	5
602	Aflatoxin B2 (ppb)	2140	< 20		1.738	0.6306	0.1253	18		602.23	5
603	Aflatoxin G1 (ppb)	0870	0.4400	0.0400				3		603.23	0
603	Aflatoxin G1 (ppb)	0910	0.4700	0.2800				3		603.21	0
603	Aflatoxin G1 (ppb)	2057	4.287	0.5393				3		603.23	0
603	Aflatoxin G1 (ppb)	2066	0.0000	0.0000				3		603.20	4
603	Aflatoxin G1 (ppb)	0032	0.0000	0.0000				3		603.23	4
603	Aflatoxin G1 (ppb)	0619	0.0000	0.0000				3		603.23	4
603	Aflatoxin G1 (ppb)	2053	< 0.1					3		603.24	5
603	Aflatoxin G1 (ppb)	0629	< 0.25					3		603.20	5
603	Aflatoxin G1 (ppb)	0682	< 0.5					3		603.24	5
603	Aflatoxin G1 (ppb)	0009	< 0.65					3		603.23	5
603	Aflatoxin G1 (ppb)	0160	< 1					3		603.23	5
603	Aflatoxin G1 (ppb)	0904	< 1					3		603.23	5
603	Aflatoxin G1 (ppb)	0964	< 1					3		603.23	5
603	Aflatoxin G1 (ppb)	2207	< 1					3		603.23	5
603	Aflatoxin G1 (ppb)	2033	< 5					3		603.23	5
603	Aflatoxin G1 (ppb)	0098	< 15					3		603.24	5
603	Aflatoxin G1 (ppb)	2140	< 20					3		603.23	5
604	Aflatoxin G2 (ppb)	0910	0.0300	0.0000				2		604.21	0
604	Aflatoxin G2 (ppb)	0870	0.0400	0.0000				2		604.23	0
604	Aflatoxin G2 (ppb)	2066	0.0000	0.0000				2		604.20	4
604	Aflatoxin G2 (ppb)	0032	0.0000	0.0000				2		604.23	4

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
604	Aflatoxin G2 (ppb)	0619	0.0000	0.0000				2	604.23	4	
604	Aflatoxin G2 (ppb)	2057	< 0.03					2	604.23	5	
604	Aflatoxin G2 (ppb)	2053	< 0.1					2	604.24	5	
604	Aflatoxin G2 (ppb)	0629	< 0.25					2	604.20	5	
604	Aflatoxin G2 (ppb)	0682	< 0.5					2	604.24	5	
604	Aflatoxin G2 (ppb)	0009	< 0.65					2	604.23	5	
604	Aflatoxin G2 (ppb)	0160	< 1					2	604.23	5	
604	Aflatoxin G2 (ppb)	0904	< 1					2	604.23	5	
604	Aflatoxin G2 (ppb)	0964	< 1					2	604.23	5	
604	Aflatoxin G2 (ppb)	2207	< 1					2	604.23	5	
604	Aflatoxin G2 (ppb)	2033	< 5					2	604.23	5	
604	Aflatoxin G2 (ppb)	0098	< 5					2	604.24	5	
604	Aflatoxin G2 (ppb)	2140	< 20					2	604.23	5	
610	Deoxynivalenol (ppb)	0619	840.7	64.73	1,879	564.0	137.9	53	-1.84	610.23	0
610	Deoxynivalenol (ppb)	0674	1,000	0.0000	1,879	564.0	137.9	53	-1.56	610.03	0
610	Deoxynivalenol (ppb)	2053	1,179	83.00	1,879	564.0	137.9	53	-1.24	610.23	0
610	Deoxynivalenol (ppb)	0563	1,197	298.0	1,879	564.0	137.9	53	-1.21	610.25	0
610	Deoxynivalenol (ppb)	0009	1,205	3.636	1,879	564.0	137.9	53	-1.19	610.23	0
610	Deoxynivalenol (ppb)	2192	1,275	30.00	1,879	564.0	137.9	53	-1.07	610.01	0
610	Deoxynivalenol (ppb)	2129	1,337	16.00	1,879	564.0	137.9	53	-0.96	610.23	0
610	Deoxynivalenol (ppb)	0032	1,350	40.00	1,879	564.0	137.9	53	-0.94	610.23	0
610	Deoxynivalenol (ppb)	0098	1,400	600.0	1,879	564.0	137.9	53	-0.85	610.20	0
610	Deoxynivalenol (ppb)	2268	1,408	110.0	1,879	564.0	137.9	53	-0.83	610.22	0
610	Deoxynivalenol (ppb)	2066	1,440	260.0	1,879	564.0	137.9	53	-0.78	610.07	0
610	Deoxynivalenol (ppb)	0870	1,520	160.0	1,879	564.0	137.9	53	-0.64	610.99	0
610	Deoxynivalenol (ppb)	2269	1,636	55.00	1,879	564.0	137.9	53	-0.43	610.98	0
610	Deoxynivalenol (ppb)	0066	1,665	130.0	1,879	564.0	137.9	53	-0.38	610.08	0
610	Deoxynivalenol (ppb)	2098	1,680	80.00	1,879	564.0	137.9	53	-0.35	610.01	0
610	Deoxynivalenol (ppb)	0013	1,750	480.0	1,879	564.0	137.9	53	-0.23	610.20	0
610	Deoxynivalenol (ppb)	2055	1,750	100.0	1,879	564.0	137.9	53	-0.23	610.23	0
610	Deoxynivalenol (ppb)	0590	1,753	145.0	1,879	564.0	137.9	53	-0.22	610.07	0
610	Deoxynivalenol (ppb)	0027	1,800	200.0	1,879	564.0	137.9	53	-0.14	610.01	0
610	Deoxynivalenol (ppb)	0208	1,840	220.0	1,879	564.0	137.9	53	-0.07	610.01	0
610	Deoxynivalenol (ppb)	0529	1,850	300.0	1,879	564.0	137.9	53	-0.05	610.01	0
610	Deoxynivalenol (ppb)	0051	1,873	108.0	1,879	564.0	137.9	53	-0.01	610.20	0
610	Deoxynivalenol (ppb)	0202	1,887	10.00	1,879	564.0	137.9	53	0.01	610.06	0
610	Deoxynivalenol (ppb)	0554	1,900	200.0	1,879	564.0	137.9	53	0.04	610.01	0
610	Deoxynivalenol (ppb)	2113	1,900	0.0000	1,879	564.0	137.9	53	0.04	610.01	0
610	Deoxynivalenol (ppb)	0868	1,923	110.0	1,879	564.0	137.9	53	0.08	610.21	0
610	Deoxynivalenol (ppb)	0010	1,925	90.00	1,879	564.0	137.9	53	0.08	610.23	0
610	Deoxynivalenol (ppb)	0904	1,931	431.4	1,879	564.0	137.9	53	0.09	610.23	0
610	Deoxynivalenol (ppb)	2140	1,935	47.00	1,879	564.0	137.9	53	0.10	610.23	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
610	Deoxynivalenol (ppb)	0142	1,950	100.0	1,879	564.0	137.9	53	0.13	610.98	0
610	Deoxynivalenol (ppb)	2050	2,000	200.0	1,879	564.0	137.9	53	0.21	610.98	0
610	Deoxynivalenol (ppb)	0037	2,000	0.0000	1,879	564.0	137.9	53	0.21	610.99	0
610	Deoxynivalenol (ppb)	0868	2,015	10.00	1,879	564.0	137.9	53	0.24	610.98	0
610	Deoxynivalenol (ppb)	0878	2,015	30.00	1,879	564.0	137.9	53	0.24	610.98	0
610	Deoxynivalenol (ppb)	2033	2,036	101.0	1,879	564.0	137.9	53	0.28	610.23	0
610	Deoxynivalenol (ppb)	2207	2,065	150.0	1,879	564.0	137.9	53	0.33	610.23	0
610	Deoxynivalenol (ppb)	0553	2,099	275.2	1,879	564.0	137.9	53	0.39	610.23	0
610	Deoxynivalenol (ppb)	0870	2,100	200.0	1,879	564.0	137.9	53	0.39	610.23	0
610	Deoxynivalenol (ppb)	0629	2,114	55.00	1,879	564.0	137.9	53	0.42	610.07	0
610	Deoxynivalenol (ppb)	0511	2,150	100.0	1,879	564.0	137.9	53	0.48	610.01	0
610	Deoxynivalenol (ppb)	2244	2,195	163.0	1,879	564.0	137.9	53	0.56	610.98	0
610	Deoxynivalenol (ppb)	0910	2,205	390.0	1,879	564.0	137.9	53	0.58	610.20	0
610	Deoxynivalenol (ppb)	0656	2,218	0.0000	1,879	564.0	137.9	53	0.60	610.07	0
610	Deoxynivalenol (ppb)	0964	2,255	15.00	1,879	564.0	137.9	53	0.67	610.23	0
610	Deoxynivalenol (ppb)	2244	2,276	11.51	1,879	564.0	137.9	53	0.70	610.20	0
610	Deoxynivalenol (ppb)	0353	2,285	10.00	1,879	564.0	137.9	53	0.72	610.23	0
610	Deoxynivalenol (ppb)	0035	2,316	231.0	1,879	564.0	137.9	53	0.77	610.08	0
610	Deoxynivalenol (ppb)	0160	2,328	30.00	1,879	564.0	137.9	53	0.80	610.23	0
610	Deoxynivalenol (ppb)	0407	2,330	15.00	1,879	564.0	137.9	53	0.80	610.07	0
610	Deoxynivalenol (ppb)	2239	2,332	20.00	1,879	564.0	137.9	53	0.80	610.98	0
610	Deoxynivalenol (ppb)	0638	2,339	481.0	1,879	564.0	137.9	53	0.81	610.07	0
610	Deoxynivalenol (ppb)	0043	2,350	110.0	1,879	564.0	137.9	53	0.84	610.23	0
610	Deoxynivalenol (ppb)	2155	2,740	237.0	1,879	564.0	137.9	53	1.53	610.23	0
610	Deoxynivalenol (ppb)	0682	1,800	0.0000	1,879	564.0	137.9	53	-3.33	610.20	3
610	Deoxynivalenol (ppb)	0042	1,900	0.2000	1,879	564.0	137.9	53	-3.33	610.01	3
610	Deoxynivalenol (ppb)	0870	2,200	0.2000	1,879	564.0	137.9	53	-3.33	610.01	3
620	Total Fumonisin (ppb)	0554	300.0	200.0	1,018	310.8	98.36	29	-2.31	620.01	0
620	Total Fumonisin (ppb)	0529	400.0	0.0000	1,018	310.8	98.36	29	-1.99	620.02	0
620	Total Fumonisin (ppb)	0619	411.9	0.0000	1,018	310.8	98.36	29	-1.95	620.23	0
620	Total Fumonisin (ppb)	0511	550.0	100.0	1,018	310.8	98.36	29	-1.51	620.01	0
620	Total Fumonisin (ppb)	0190	600.0	0.0000	1,018	310.8	98.36	29	-1.35	620.11	0
620	Total Fumonisin (ppb)	2066	645.0	50.00	1,018	310.8	98.36	29	-1.20	620.09	0
620	Total Fumonisin (ppb)	0035	727.5	55.00	1,018	310.8	98.36	29	-0.94	620.09	0
620	Total Fumonisin (ppb)	0202	764.0	188.0	1,018	310.8	98.36	29	-0.82	620.09	0
620	Total Fumonisin (ppb)	0098	800.0	0.0000	1,018	310.8	98.36	29	-0.70	620.11	0
620	Total Fumonisin (ppb)	0870	850.0	100.0	1,018	310.8	98.36	29	-0.54	620.02	0
620	Total Fumonisin (ppb)	2098	890.0	20.00	1,018	310.8	98.36	29	-0.41	620.02	0
620	Total Fumonisin (ppb)	2269	915.5	173.0	1,018	310.8	98.36	29	-0.33	620.98	0
620	Total Fumonisin (ppb)	0027	945.0	310.0	1,018	310.8	98.36	29	-0.24	620.01	0
620	Total Fumonisin (ppb)	0674	950.0	0.0000	1,018	310.8	98.36	29	-0.22	620.05	0
620	Total Fumonisin (ppb)	0868	1,015	30.00	1,018	310.8	98.36	29	-0.01	620.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
620	Total Fumonisin (ppb)	2244	1,064	87.41	1,018	310.8	98.36	29	0.15	620.23	0
620	Total Fumonisin (ppb)	2226	1,100	34.00	1,018	310.8	98.36	29	0.26	620.09	0
620	Total Fumonisin (ppb)	0010	1,100	0.0000	1,018	310.8	98.36	29	0.26	620.23	0
620	Total Fumonisin (ppb)	0407	1,113	122.0	1,018	310.8	98.36	29	0.30	620.09	0
620	Total Fumonisin (ppb)	0066	1,200	0.0000	1,018	310.8	98.36	29	0.58	620.09	0
620	Total Fumonisin (ppb)	0160	1,250	100.0	1,018	310.8	98.36	29	0.75	620.23	0
620	Total Fumonisin (ppb)	0656	1,265	0.0000	1,018	310.8	98.36	29	0.79	620.09	0
620	Total Fumonisin (ppb)	2239	1,380	63.00	1,018	310.8	98.36	29	1.16	620.98	0
620	Total Fumonisin (ppb)	0208	1,400	400.0	1,018	310.8	98.36	29	1.23	620.20	0
620	Total Fumonisin (ppb)	0032	1,430	60.00	1,018	310.8	98.36	29	1.32	620.23	0
620	Total Fumonisin (ppb)	0142	1,500	200.0	1,018	310.8	98.36	29	1.55	620.98	0
620	Total Fumonisin (ppb)	0353	1,575	10.00	1,018	310.8	98.36	29	1.79	620.23	0
620	Total Fumonisin (ppb)	2033	2,184	150.0	1,018	310.8	98.36	29	3.75	620.23	0
620	Total Fumonisin (ppb)	2268	3,900	400.0	1,018	310.8	98.36	29	9.27	620.11	0
620	Total Fumonisin (ppb)	0682	1,500	0.0000	1,018	310.8	98.36	29	-3.27	620.11	3
620	Total Fumonisin (ppb)	0042	< 2		1,018	310.8	98.36	29		620.01	5
620	Total Fumonisin (ppb)	2113	< 1000		1,018	310.8	98.36	29		620.01	5
621	Fumonsin B1 (ppb)	0619	274.1	50.17	898.8	275.3	81.33	20	-2.27	621.23	0
621	Fumonsin B1 (ppb)	2053	491.5	33.00	898.8	275.3	81.33	20	-1.48	621.22	0
621	Fumonsin B1 (ppb)	2155	492.5	365.0	898.8	275.3	81.33	20	-1.48	621.23	0
621	Fumonsin B1 (ppb)	0098	600.0	0.0000	898.8	275.3	81.33	20	-1.09	621.21	0
621	Fumonsin B1 (ppb)	0964	650.4	7.400	898.8	275.3	81.33	20	-0.90	621.23	0
621	Fumonsin B1 (ppb)	0043	705.0	10.00	898.8	275.3	81.33	20	-0.70	621.23	0
621	Fumonsin B1 (ppb)	2244	740.6	48.59	898.8	275.3	81.33	20	-0.57	621.20	0
621	Fumonsin B1 (ppb)	0553	774.2	32.70	898.8	275.3	81.33	20	-0.45	621.23	0
621	Fumonsin B1 (ppb)	0870	790.0	100.0	898.8	275.3	81.33	20	-0.40	621.23	0
621	Fumonsin B1 (ppb)	0009	854.8	76.96	898.8	275.3	81.33	20	-0.16	621.23	0
621	Fumonsin B1 (ppb)	0160	938.0	48.00	898.8	275.3	81.33	20	0.14	621.23	0
621	Fumonsin B1 (ppb)	0910	975.0	52.00	898.8	275.3	81.33	20	0.28	621.21	0
621	Fumonsin B1 (ppb)	0032	1,002	16.00	898.8	275.3	81.33	20	0.37	621.23	0
621	Fumonsin B1 (ppb)	0904	1,078	89.78	898.8	275.3	81.33	20	0.65	621.23	0
621	Fumonsin B1 (ppb)	0353	1,103	20.00	898.8	275.3	81.33	20	0.74	621.23	0
621	Fumonsin B1 (ppb)	2140	1,122	12.00	898.8	275.3	81.33	20	0.81	621.23	0
621	Fumonsin B1 (ppb)	2207	1,150	100.0	898.8	275.3	81.33	20	0.91	621.23	0
621	Fumonsin B1 (ppb)	2055	1,300	400.0	898.8	275.3	81.33	20	1.46	621.23	0
621	Fumonsin B1 (ppb)	0563	1,553	75.00	898.8	275.3	81.33	20	2.37	621.20	0
621	Fumonsin B1 (ppb)	2033	1,625	90.00	898.8	275.3	81.33	20	2.64	621.23	0
621	Fumonsin B1 (ppb)	0682	1,100	0.0000	898.8	275.3	81.33	20	-3.26	621.20	3
622	Fumonisin B2 (ppb)	0619	90.24	44.89	317.2	99.92	25.20	18	-2.27	622.23	0
622	Fumonisin B2 (ppb)	2053	186.0	12.00	317.2	99.92	25.20	18	-1.31	622.22	0
622	Fumonisin B2 (ppb)	0098	200.0	0.0000	317.2	99.92	25.20	18	-1.17	622.21	0
622	Fumonisin B2 (ppb)	2207	200.0	0.0000	317.2	99.92	25.20	18	-1.17	622.23	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
622	Fumonisin B2 (ppb)	0870	260.0	40.00	317.2	99.92	25.20	18	-0.57	622.23	0
622	Fumonisin B2 (ppb)	0160	266.5	11.00	317.2	99.92	25.20	18	-0.51	622.23	0
622	Fumonisin B2 (ppb)	0910	287.0	2.000	317.2	99.92	25.20	18	-0.30	622.21	0
622	Fumonisin B2 (ppb)	0032	317.0	36.00	317.2	99.92	25.20	18	0.00	622.23	0
622	Fumonisin B2 (ppb)	2244	323.7	38.82	317.2	99.92	25.20	18	0.07	622.20	0
622	Fumonisin B2 (ppb)	2140	330.0	50.00	317.2	99.92	25.20	18	0.13	622.23	0
622	Fumonisin B2 (ppb)	0964	338.8	4.800	317.2	99.92	25.20	18	0.22	622.23	0
622	Fumonisin B2 (ppb)	0043	340.7	10.70	317.2	99.92	25.20	18	0.23	622.23	0
622	Fumonisin B2 (ppb)	0009	340.7	66.83	317.2	99.92	25.20	18	0.24	622.23	0
622	Fumonisin B2 (ppb)	0553	371.0	41.00	317.2	99.92	25.20	18	0.54	622.23	0
622	Fumonisin B2 (ppb)	2055	400.0	0.0000	317.2	99.92	25.20	18	0.83	622.23	0
622	Fumonisin B2 (ppb)	2033	441.5	53.00	317.2	99.92	25.20	18	1.24	622.23	0
622	Fumonisin B2 (ppb)	0353	472.0	20.00	317.2	99.92	25.20	18	1.55	622.23	0
622	Fumonisin B2 (ppb)	0904	482.4	22.63	317.2	99.92	25.20	18	1.65	622.23	0
622	Fumonisin B2 (ppb)	2155	400.5	455.0	317.2	99.92	25.20	18	0.83	622.23	1
622	Fumonisin B2 (ppb)	0682	0.4000	0.0000	317.2	99.92	25.20	18	-3.17	622.20	3
623	Fumonisin B3 (ppb)	0870	0.0900	0.0200	79.85	26.11	9.441	7	-3.05	623.23	0
623	Fumonisin B3 (ppb)	2155	54.00	22.00	79.85	26.11	9.441	7	-0.99	623.23	0
623	Fumonisin B3 (ppb)	2053	59.05	6.100	79.85	26.11	9.441	7	-0.80	623.22	0
623	Fumonisin B3 (ppb)	0553	91.05	11.31	79.85	26.11	9.441	7	0.43	623.23	0
623	Fumonisin B3 (ppb)	0032	108.0	8.000	79.85	26.11	9.441	7	1.08	623.23	0
623	Fumonisin B3 (ppb)	0009	116.0	11.66	79.85	26.11	9.441	7	1.38	623.23	0
623	Fumonisin B3 (ppb)	2033	117.5	7.000	79.85	26.11	9.441	7	1.44	623.23	0
623	Fumonisin B3 (ppb)	0160	< 100		79.85	26.11	9.441	7		623.23	5
630	Ochratoxin A (ppb)	0529	8.700	1.400	41.26	13.74	4.544	39	-2.37	630.01	0
630	Ochratoxin A (ppb)	2098	9.750	0.5000	41.26	13.74	4.544	39	-2.29	630.01	0
630	Ochratoxin A (ppb)	0619	14.15	6.100	41.26	13.74	4.544	39	-1.97	630.23	0
630	Ochratoxin A (ppb)	0674	16.40	0.0000	41.26	13.74	4.544	39	-1.81	630.20	0
630	Ochratoxin A (ppb)	0682	21.30	0.0000	41.26	13.74	4.544	39	-1.45	630.20	0
630	Ochratoxin A (ppb)	0656	21.39	0.0000	41.26	13.74	4.544	39	-1.45	630.05	0
630	Ochratoxin A (ppb)	2066	25.10	0.8000	41.26	13.74	4.544	39	-1.18	630.04	0
630	Ochratoxin A (ppb)	2066	25.25	1.100	41.26	13.74	4.544	39	-1.17	630.20	0
630	Ochratoxin A (ppb)	2269	27.92	6.650	41.26	13.74	4.544	39	-0.97	630.98	0
630	Ochratoxin A (ppb)	0629	28.00	4.000	41.26	13.74	4.544	39	-0.97	630.04	0
630	Ochratoxin A (ppb)	0563	30.50	1.000	41.26	13.74	4.544	39	-0.78	630.20	0
630	Ochratoxin A (ppb)	2053	32.00	4.600	41.26	13.74	4.544	39	-0.67	630.23	0
630	Ochratoxin A (ppb)	0870	32.36	2.000	41.26	13.74	4.544	39	-0.65	630.23	0
630	Ochratoxin A (ppb)	0066	32.55	2.100	41.26	13.74	4.544	39	-0.63	630.05	0
630	Ochratoxin A (ppb)	2192	34.15	2.300	41.26	13.74	4.544	39	-0.52	630.20	0
630	Ochratoxin A (ppb)	0407	34.81	2.410	41.26	13.74	4.544	39	-0.47	630.04	0
630	Ochratoxin A (ppb)	2244	36.19	1.367	41.26	13.74	4.544	39	-0.37	630.20	0
630	Ochratoxin A (ppb)	0868	36.66	0.5500	41.26	13.74	4.544	39	-0.34	630.03	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
630	Ochratoxin A (ppb)	2244	36.89	1.390	41.26	13.74	4.544	39	-0.32	630.98	0
630	Ochratoxin A (ppb)	0878	42.95	5.540	41.26	13.74	4.544	39	0.12	630.98	0
630	Ochratoxin A (ppb)	0870	44.70	1.000	41.26	13.74	4.544	39	0.25	630.98	0
630	Ochratoxin A (ppb)	2239	45.55	1.900	41.26	13.74	4.544	39	0.31	630.98	0
630	Ochratoxin A (ppb)	2226	46.41	10.62	41.26	13.74	4.544	39	0.37	630.05	0
630	Ochratoxin A (ppb)	0353	47.00	2.000	41.26	13.74	4.544	39	0.42	630.23	0
630	Ochratoxin A (ppb)	0904	47.33	5.800	41.26	13.74	4.544	39	0.44	630.23	0
630	Ochratoxin A (ppb)	0010	51.40	12.40	41.26	13.74	4.544	39	0.74	630.23	0
630	Ochratoxin A (ppb)	2207	53.00	6.000	41.26	13.74	4.544	39	0.85	630.23	0
630	Ochratoxin A (ppb)	2055	53.80	1.800	41.26	13.74	4.544	39	0.91	630.23	0
630	Ochratoxin A (ppb)	0043	53.97	0.2700	41.26	13.74	4.544	39	0.92	630.23	0
630	Ochratoxin A (ppb)	0051	54.00	6.000	41.26	13.74	4.544	39	0.93	630.20	0
630	Ochratoxin A (ppb)	0098	55.25	0.9000	41.26	13.74	4.544	39	1.02	630.20	0
630	Ochratoxin A (ppb)	0032	56.10	5.800	41.26	13.74	4.544	39	1.08	630.23	0
630	Ochratoxin A (ppb)	0160	56.22	10.45	41.26	13.74	4.544	39	1.09	630.23	0
630	Ochratoxin A (ppb)	0910	60.00	4.600	41.26	13.74	4.544	39	1.36	630.21	0
630	Ochratoxin A (ppb)	2140	60.50	9.000	41.26	13.74	4.544	39	1.40	630.23	0
630	Ochratoxin A (ppb)	0964	64.90	13.00	41.26	13.74	4.544	39	1.72	630.23	0
630	Ochratoxin A (ppb)	2033	65.50	1.000	41.26	13.74	4.544	39	1.76	630.23	0
630	Ochratoxin A (ppb)	0009	70.39	6.013	41.26	13.74	4.544	39	2.12	630.23	0
630	Ochratoxin A (ppb)	0553	72.57	34.86	41.26	13.74	4.544	39	2.28	630.23	0
640	T-2 Toxin (ppb)	0619	34.55	21.50	60.57	19.96	4.876	24	-1.30	640.23	0
640	T-2 Toxin (ppb)	0910	39.00	6.000	60.57	19.96	4.876	24	-1.08	640.21	0
640	T-2 Toxin (ppb)	0066	43.00	1.400	60.57	19.96	4.876	24	-0.88	640.03	0
640	T-2 Toxin (ppb)	0553	43.07	4.670	60.57	19.96	4.876	24	-0.88	640.23	0
640	T-2 Toxin (ppb)	2033	45.00	10.00	60.57	19.96	4.876	24	-0.78	640.23	0
640	T-2 Toxin (ppb)	0878	46.15	2.900	60.57	19.96	4.876	24	-0.72	640.98	0
640	T-2 Toxin (ppb)	0682	47.00	0.0000	60.57	19.96	4.876	24	-0.68	640.23	0
640	T-2 Toxin (ppb)	2053	50.35	3.300	60.57	19.96	4.876	24	-0.51	640.23	0
640	T-2 Toxin (ppb)	2055	52.55	6.500	60.57	19.96	4.876	24	-0.40	640.23	0
640	T-2 Toxin (ppb)	2129	54.20	2.000	60.57	19.96	4.876	24	-0.32	640.23	0
640	T-2 Toxin (ppb)	0868	56.55	3.080	60.57	19.96	4.876	24	-0.20	640.03	0
640	T-2 Toxin (ppb)	0870	56.65	0.9800	60.57	19.96	4.876	24	-0.20	640.23	0
640	T-2 Toxin (ppb)	2140	57.00	2.000	60.57	19.96	4.876	24	-0.18	640.23	0
640	T-2 Toxin (ppb)	0904	57.35	2.300	60.57	19.96	4.876	24	-0.16	640.23	0
640	T-2 Toxin (ppb)	0353	60.00	2.000	60.57	19.96	4.876	24	-0.03	640.23	0
640	T-2 Toxin (ppb)	2244	60.20	7.420	60.57	19.96	4.876	24	-0.02	640.23	0
640	T-2 Toxin (ppb)	0407	62.67	4.020	60.57	19.96	4.876	24	0.11	640.04	0
640	T-2 Toxin (ppb)	0638	70.00	12.00	60.57	19.96	4.876	24	0.47	640.05	0
640	T-2 Toxin (ppb)	0674	71.00	0.0000	60.57	19.96	4.876	24	0.52	640.98	0
640	T-2 Toxin (ppb)	2239	86.10	1.000	60.57	19.96	4.876	24	1.28	640.98	0
640	T-2 Toxin (ppb)	2269	90.28	11.85	60.57	19.96	4.876	24	1.49	640.98	0



Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
640	T-2 Toxin (ppb)	2207	95.00	10.00	60.57	19.96	4.876	24	1.73	640.23	0
640	T-2 Toxin (ppb)	0656	96.73	0.0000	60.57	19.96	4.876	24	1.81	640.05	0
640	T-2 Toxin (ppb)	0964	97.25	2.100	60.57	19.96	4.876	24	1.84	640.23	0
640	T-2 Toxin (ppb)	0160	< 100		60.57	19.96	4.876	24		640.23	5
641	HT-2 Toxin (ppb)	0619	25.70	11.00	43.81	14.56	5.850	10	-1.24	641.23	0
641	HT-2 Toxin (ppb)	0870	26.76	2.200	43.81	14.56	5.850	10	-1.17	641.23	0
641	HT-2 Toxin (ppb)	0910	30.00	4.000	43.81	14.56	5.850	10	-0.95	641.21	0
641	HT-2 Toxin (ppb)	2033	33.95	2.100	43.81	14.56	5.850	10	-0.68	641.23	0
641	HT-2 Toxin (ppb)	0353	37.00	2.000	43.81	14.56	5.850	10	-0.47	641.23	0
641	HT-2 Toxin (ppb)	2053	41.00	5.200	43.81	14.56	5.850	10	-0.19	641.23	0
641	HT-2 Toxin (ppb)	2140	48.50	3.000	43.81	14.56	5.850	10	0.32	641.23	0
641	HT-2 Toxin (ppb)	2055	52.65	5.300	43.81	14.56	5.850	10	0.61	641.23	0
641	HT-2 Toxin (ppb)	2207	70.00	20.00	43.81	14.56	5.850	10	1.80	641.23	0
641	HT-2 Toxin (ppb)	0964	83.05	3.700	43.81	14.56	5.850	10	2.69	641.23	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0142	61.25	6.100	102.1	33.16	13.17	11	-1.23	642.98	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2098	79.55	0.3000	102.1	33.16	13.17	11	-0.68	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0407	83.52	6.880	102.1	33.16	13.17	11	-0.56	642.99	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0629	87.50	1.000	102.1	33.16	13.17	11	-0.44	642.98	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0590	89.00	34.00	102.1	33.16	13.17	11	-0.39	642.98	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0870	95.65	4.900	102.1	33.16	13.17	11	-0.19	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0037	114.0	3.700	102.1	33.16	13.17	11	0.36	642.99	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2244	120.4	14.84	102.1	33.16	13.17	11	0.55	642.98	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0529	123.3	4.750	102.1	33.16	13.17	11	0.64	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2269	124.6	16.35	102.1	33.16	13.17	11	0.68	642.98	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2192	236.0	52.00	102.1	33.16	13.17	11	4.04	642.21	0
650	Zearalenone (ppb)	2268	270.0	40.00	3,766	1,110	242.3	45	-3.15	650.99	0
650	Zearalenone (ppb)	0674	400.0	0.0000	3,766	1,110	242.3	45	-3.03	650.02	0
650	Zearalenone (ppb)	0027	961.3	129.6	3,766	1,110	242.3	45	-2.53	650.01	0
650	Zearalenone (ppb)	0682	980.0	0.0000	3,766	1,110	242.3	45	-2.51	650.20	0
650	Zearalenone (ppb)	0035	1,018	64.96	3,766	1,110	242.3	45	-2.48	650.07	0
650	Zearalenone (ppb)	0563	1,055	40.00	3,766	1,110	242.3	45	-2.44	650.20	0
650	Zearalenone (ppb)	2066	1,070	57.00	3,766	1,110	242.3	45	-2.43	650.06	0
650	Zearalenone (ppb)	0142	1,647	43.70	3,766	1,110	242.3	45	-1.91	650.98	0
650	Zearalenone (ppb)	0202	2,029	33.00	3,766	1,110	242.3	45	-1.57	650.06	0
650	Zearalenone (ppb)	0619	2,065	61.39	3,766	1,110	242.3	45	-1.53	650.24	0
650	Zearalenone (ppb)	0656	2,184	0.0000	3,766	1,110	242.3	45	-1.43	650.06	0
650	Zearalenone (ppb)	2113	2,356	388.0	3,766	1,110	242.3	45	-1.27	650.01	0
650	Zearalenone (ppb)	0066	2,510	117.3	3,766	1,110	242.3	45	-1.13	650.07	0
650	Zearalenone (ppb)	0037	2,668	1.000	3,766	1,110	242.3	45	-0.99	650.99	0
650	Zearalenone (ppb)	0098	2,710	420.0	3,766	1,110	242.3	45	-0.95	650.20	0
650	Zearalenone (ppb)	0870	2,817	40.00	3,766	1,110	242.3	45	-0.86	650.01	0
650	Zearalenone (ppb)	0870	2,833	154.0	3,766	1,110	242.3	45	-0.84	650.24	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
650	Zearalenone (ppb)	2269	3,570	226.6	3,766	1,110	242.3	45	-0.18	650.98	0
650	Zearalenone (ppb)	2192	3,715	191.0	3,766	1,110	242.3	45	-0.05	650.20	0
650	Zearalenone (ppb)	0529	3,852	4.000	3,766	1,110	242.3	45	0.08	650.01	0
650	Zearalenone (ppb)	0407	4,026	40.53	3,766	1,110	242.3	45	0.23	650.07	0
650	Zearalenone (ppb)	2098	4,030	696.2	3,766	1,110	242.3	45	0.24	650.01	0
650	Zearalenone (ppb)	2207	4,160	340.0	3,766	1,110	242.3	45	0.35	650.24	0
650	Zearalenone (ppb)	2053	4,196	643.0	3,766	1,110	242.3	45	0.39	650.24	0
650	Zearalenone (ppb)	0910	4,278	485.0	3,766	1,110	242.3	45	0.46	650.20	0
650	Zearalenone (ppb)	0878	4,370	4.000	3,766	1,110	242.3	45	0.54	650.98	0
650	Zearalenone (ppb)	0553	4,603	1,818	3,766	1,110	242.3	45	0.75	650.24	0
650	Zearalenone (ppb)	2244	4,609	101.6	3,766	1,110	242.3	45	0.76	650.20	0
650	Zearalenone (ppb)	2244	4,621	24.86	3,766	1,110	242.3	45	0.77	650.98	0
650	Zearalenone (ppb)	0868	4,660	175.0	3,766	1,110	242.3	45	0.80	650.21	0
650	Zearalenone (ppb)	2226	4,786	542.0	3,766	1,110	242.3	45	0.92	650.06	0
650	Zearalenone (ppb)	2129	4,802	131.0	3,766	1,110	242.3	45	0.93	650.24	0
650	Zearalenone (ppb)	0353	4,867	20.00	3,766	1,110	242.3	45	0.99	650.23	0
650	Zearalenone (ppb)	2055	5,100	600.0	3,766	1,110	242.3	45	1.20	650.23	0
650	Zearalenone (ppb)	0160	5,235	365.0	3,766	1,110	242.3	45	1.32	650.23	0
650	Zearalenone (ppb)	0964	5,388	134.9	3,766	1,110	242.3	45	1.46	650.24	0
650	Zearalenone (ppb)	0868	5,480	44.00	3,766	1,110	242.3	45	1.54	650.98	0
650	Zearalenone (ppb)	2140	5,636	538.0	3,766	1,110	242.3	45	1.68	650.24	0
650	Zearalenone (ppb)	0043	5,837	20.00	3,766	1,110	242.3	45	1.87	650.24	0
650	Zearalenone (ppb)	0010	5,900	200.0	3,766	1,110	242.3	45	1.92	650.24	0
650	Zearalenone (ppb)	0032	5,905	110.0	3,766	1,110	242.3	45	1.93	650.24	0
650	Zearalenone (ppb)	2033	6,000	200.0	3,766	1,110	242.3	45	2.01	650.24	0
650	Zearalenone (ppb)	2239	6,239	118.0	3,766	1,110	242.3	45	2.23	650.98	0
650	Zearalenone (ppb)	0051	6,435	400.0	3,766	1,110	242.3	45	2.41	650.20	0
650	Zearalenone (ppb)	2155	7,543	1,142	3,766	1,110	242.3	45	3.40	650.24	0

Notes: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2 and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = Rejected for duplicates too far apart, 2 = Rejected as outlier, 5 = A reporting limit and 4 = zeros submitted as values. Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). Flag 3 indicates not used in statistics.