



AAFCO
Proficiency Testing Program



Animal Feed Scheme

Rabbit Feed, Residue

Test Material Code # 202127

Labs Reporting: 164

Analytes Reported 108

Issue Date : 08/31/2021

Analyte Proficiency Testing Report

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
000	Urea (%)	0278	0.3000	0.0000				1		000.02	0
001	Loss on Drying (%)	2259	6.988	0.1266	8.350	0.3266	0.1206	67	-4.17	001.07	0
001	Loss on Drying (%)	2322	7.510	0.0600	8.350	0.3266	0.1206	67	-2.57	001.07	0
001	Loss on Drying (%)	0038	7.675	0.0500	8.350	0.3266	0.1206	67	-2.07	001.07	0
001	Loss on Drying (%)	0353	7.705	0.0500	8.350	0.3266	0.1206	67	-1.97	001.07	0
001	Loss on Drying (%)	0345	7.750	0.1600	8.350	0.3266	0.1206	67	-1.84	001.07	0
001	Loss on Drying (%)	0034	7.780	0.1200	8.350	0.3266	0.1206	67	-1.74	001.07	0
001	Loss on Drying (%)	2155	7.785	0.3500	8.350	0.3266	0.1206	67	-1.73	001.99	0
001	Loss on Drying (%)	0643	7.840	0.0400	8.350	0.3266	0.1206	67	-1.56	001.07	0
001	Loss on Drying (%)	0169	7.955	0.0100	8.350	0.3266	0.1206	67	-1.21	001.00	0
001	Loss on Drying (%)	0366	8.000	0.6000	8.350	0.3266	0.1206	67	-1.07	001.07	0
001	Loss on Drying (%)	0910	8.000	0.0000	8.350	0.3266	0.1206	67	-1.07	001.99	0
001	Loss on Drying (%)	0066	8.045	0.1300	8.350	0.3266	0.1206	67	-0.93	001.07	0
001	Loss on Drying (%)	0226	8.060	0.0400	8.350	0.3266	0.1206	67	-0.89	001.07	0
001	Loss on Drying (%)	0504	8.065	0.0900	8.350	0.3266	0.1206	67	-0.87	001.00	0
001	Loss on Drying (%)	0027	8.105	0.1320	8.350	0.3266	0.1206	67	-0.75	001.07	0
001	Loss on Drying (%)	0027	8.124	0.0500	8.350	0.3266	0.1206	67	-0.69	001.00	0
001	Loss on Drying (%)	2081	8.145	0.3700	8.350	0.3266	0.1206	67	-0.63	001.99	0
001	Loss on Drying (%)	0610	8.165	0.0100	8.350	0.3266	0.1206	67	-0.56	001.99	0
001	Loss on Drying (%)	0083	8.170	0.0400	8.350	0.3266	0.1206	67	-0.55	001.07	0
001	Loss on Drying (%)	0510	8.200	0.0000	8.350	0.3266	0.1206	67	-0.46	001.99	0
001	Loss on Drying (%)	0357	8.204	0.0449	8.350	0.3266	0.1206	67	-0.44	001.99	0
001	Loss on Drying (%)	0948	8.235	0.0300	8.350	0.3266	0.1206	67	-0.35	001.99	0
001	Loss on Drying (%)	0581	8.250	0.1000	8.350	0.3266	0.1206	67	-0.30	001.07	0
001	Loss on Drying (%)	0676	8.250	0.0400	8.350	0.3266	0.1206	67	-0.30	001.99	0
001	Loss on Drying (%)	0407	8.252	0.2270	8.350	0.3266	0.1206	67	-0.30	001.07	0
001	Loss on Drying (%)	2268	8.270	0.0800	8.350	0.3266	0.1206	67	-0.24	001.99	0
001	Loss on Drying (%)	0037	8.285	0.1100	8.350	0.3266	0.1206	67	-0.20	001.99	0
001	Loss on Drying (%)	0309	8.304	0.2190	8.350	0.3266	0.1206	67	-0.14	001.00	0
001	Loss on Drying (%)	0693	8.313	0.0000	8.350	0.3266	0.1206	67	-0.11	001.07	0
001	Loss on Drying (%)	0843	8.345	0.2700	8.350	0.3266	0.1206	67	-0.01	001.03	0
001	Loss on Drying (%)	0843	8.345	0.2700	8.350	0.3266	0.1206	67	-0.01	001.07	0
001	Loss on Drying (%)	0049	8.365	0.4300	8.350	0.3266	0.1206	67	0.05	001.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
001	Loss on Drying (%)	0028	8.390	0.1400	8.350	0.3266	0.1206	67	0.12	001.07	0
001	Loss on Drying (%)	0035	8.390	0.0400	8.350	0.3266	0.1206	67	0.12	001.07	0
001	Loss on Drying (%)	0089	8.390	0.0400	8.350	0.3266	0.1206	67	0.12	001.07	0
001	Loss on Drying (%)	0876	8.400	0.2000	8.350	0.3266	0.1206	67	0.15	001.07	0
001	Loss on Drying (%)	2109	8.400	0.0200	8.350	0.3266	0.1206	67	0.15	001.07	0
001	Loss on Drying (%)	2009	8.414	0.0159	8.350	0.3266	0.1206	67	0.20	001.07	0
001	Loss on Drying (%)	0142	8.415	0.0100	8.350	0.3266	0.1206	67	0.20	001.07	0
001	Loss on Drying (%)	2192	8.435	0.0100	8.350	0.3266	0.1206	67	0.26	001.07	0
001	Loss on Drying (%)	2059	8.450	0.0200	8.350	0.3266	0.1206	67	0.31	001.07	0
001	Loss on Drying (%)	2190	8.465	0.0500	8.350	0.3266	0.1206	67	0.35	001.07	0
001	Loss on Drying (%)	0171	8.475	0.0300	8.350	0.3266	0.1206	67	0.38	001.07	0
001	Loss on Drying (%)	0033	8.485	0.0500	8.350	0.3266	0.1206	67	0.41	001.07	0
001	Loss on Drying (%)	0297	8.485	0.1300	8.350	0.3266	0.1206	67	0.41	001.07	0
001	Loss on Drying (%)	0675	8.495	0.0300	8.350	0.3266	0.1206	67	0.45	001.07	0
001	Loss on Drying (%)	0278	8.500	0.0600	8.350	0.3266	0.1206	67	0.46	001.07	0
001	Loss on Drying (%)	0505	8.500	0.0400	8.350	0.3266	0.1206	67	0.46	001.07	0
001	Loss on Drying (%)	0612	8.505	0.2300	8.350	0.3266	0.1206	67	0.48	001.99	0
001	Loss on Drying (%)	0003	8.540	0.1400	8.350	0.3266	0.1206	67	0.58	001.07	0
001	Loss on Drying (%)	2319	8.550	0.1000	8.350	0.3266	0.1206	67	0.61	001.99	0
001	Loss on Drying (%)	0644	8.569	0.0230	8.350	0.3266	0.1206	67	0.67	001.05	0
001	Loss on Drying (%)	2073	8.575	0.2100	8.350	0.3266	0.1206	67	0.69	001.00	0
001	Loss on Drying (%)	0571	8.580	0.0800	8.350	0.3266	0.1206	67	0.71	001.07	0
001	Loss on Drying (%)	0619	8.585	0.0100	8.350	0.3266	0.1206	67	0.72	001.03	0
001	Loss on Drying (%)	0045	8.600	0.4000	8.350	0.3266	0.1206	67	0.77	001.07	0
001	Loss on Drying (%)	0148	8.605	0.0500	8.350	0.3266	0.1206	67	0.78	001.07	0
001	Loss on Drying (%)	0689	8.650	0.1000	8.350	0.3266	0.1206	67	0.92	001.07	0
001	Loss on Drying (%)	0853	8.700	0.0000	8.350	0.3266	0.1206	67	1.07	001.07	0
001	Loss on Drying (%)	2345	8.760	0.0400	8.350	0.3266	0.1206	67	1.26	001.07	0
001	Loss on Drying (%)	0019	8.780	0.2000	8.350	0.3266	0.1206	67	1.32	001.99	0
001	Loss on Drying (%)	0098	8.805	0.2100	8.350	0.3266	0.1206	67	1.39	001.07	0
001	Loss on Drying (%)	0682	9.020	0.0000	8.350	0.3266	0.1206	67	2.05	001.07	0
001	Loss on Drying (%)	0852	9.100	0.2000	8.350	0.3266	0.1206	67	2.30	001.99	0
001	Loss on Drying (%)	0720	9.260	0.8800	8.350	0.3266	0.1206	67	2.79	001.99	0
001	Loss on Drying (%)	0536	9.510	0.0400	8.350	0.3266	0.1206	67	3.55	001.99	0
001	Loss on Drying (%)	2328	10.44	0.0300	8.350	0.3266	0.1206	67	6.38	001.99	0
001	Loss on Drying (%)	0019	6.450	1.900	8.350	0.3266	0.1206	67	-5.82	001.07	1
002	Protein, Crude (%)	0620	15.61	0.0000	17.26	0.3595	0.1709	156	-4.58	002.05	0
002	Protein, Crude (%)	2268	16.16	0.1300	17.26	0.3595	0.1709	156	-3.08	002.06	0
002	Protein, Crude (%)	0520	16.34	0.0800	17.26	0.3595	0.1709	156	-2.56	002.06	0
002	Protein, Crude (%)	0953	16.39	0.4600	17.26	0.3595	0.1709	156	-2.43	002.06	0
002	Protein, Crude (%)	2328	16.47	0.3800	17.26	0.3595	0.1709	156	-2.20	002.00	0
002	Protein, Crude (%)	0529	16.54	0.2900	17.26	0.3595	0.1709	156	-2.02	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0536	16.54	0.0600	17.26	0.3595	0.1709	156	-2.01	002.05	0
002	Protein, Crude (%)	0539	16.58	0.3100	17.26	0.3595	0.1709	156	-1.91	002.06	0
002	Protein, Crude (%)	2006	16.60	0.0000	17.26	0.3595	0.1709	156	-1.84	002.05	0
002	Protein, Crude (%)	0504	16.65	0.5400	17.26	0.3595	0.1709	156	-1.70	002.04	0
002	Protein, Crude (%)	0939	16.66	0.4600	17.26	0.3595	0.1709	156	-1.67	002.05	0
002	Protein, Crude (%)	2196	16.67	0.0000	17.26	0.3595	0.1709	156	-1.65	002.01	0
002	Protein, Crude (%)	0939	16.69	0.5900	17.26	0.3595	0.1709	156	-1.60	002.06	0
002	Protein, Crude (%)	0689	16.75	0.1000	17.26	0.3595	0.1709	156	-1.42	002.05	0
002	Protein, Crude (%)	0144	16.75	0.2600	17.26	0.3595	0.1709	156	-1.42	002.06	0
002	Protein, Crude (%)	0876	16.80	0.2000	17.26	0.3595	0.1709	156	-1.29	002.06	0
002	Protein, Crude (%)	0598	16.81	0.4900	17.26	0.3595	0.1709	156	-1.27	002.06	0
002	Protein, Crude (%)	0676	16.82	0.2500	17.26	0.3595	0.1709	156	-1.24	002.06	0
002	Protein, Crude (%)	0921	16.85	0.0000	17.26	0.3595	0.1709	156	-1.15	002.05	0
002	Protein, Crude (%)	0870	16.85	0.3560	17.26	0.3595	0.1709	156	-1.14	002.01	0
002	Protein, Crude (%)	2325	16.88	0.0100	17.26	0.3595	0.1709	156	-1.08	002.01	0
002	Protein, Crude (%)	2184	16.89	0.3200	17.26	0.3595	0.1709	156	-1.03	002.06	0
002	Protein, Crude (%)	0505	16.90	0.1300	17.26	0.3595	0.1709	156	-1.02	002.06	0
002	Protein, Crude (%)	0885	16.90	0.0110	17.26	0.3595	0.1709	156	-1.02	002.05	0
002	Protein, Crude (%)	2073	16.91	0.0500	17.26	0.3595	0.1709	156	-0.99	002.05	0
002	Protein, Crude (%)	0644	16.91	0.2650	17.26	0.3595	0.1709	156	-0.99	002.06	0
002	Protein, Crude (%)	0357	16.91	0.0625	17.26	0.3595	0.1709	156	-0.99	002.06	0
002	Protein, Crude (%)	0675	16.92	0.0800	17.26	0.3595	0.1709	156	-0.95	002.05	0
002	Protein, Crude (%)	0948	16.93	0.0100	17.26	0.3595	0.1709	156	-0.94	002.05	0
002	Protein, Crude (%)	2009	16.98	0.0051	17.26	0.3595	0.1709	156	-0.79	002.05	0
002	Protein, Crude (%)	0036	16.99	0.0414	17.26	0.3595	0.1709	156	-0.76	002.02	0
002	Protein, Crude (%)	0610	17.00	0.2000	17.26	0.3595	0.1709	156	-0.73	002.01	0
002	Protein, Crude (%)	0047	17.00	0.2000	17.26	0.3595	0.1709	156	-0.73	002.06	0
002	Protein, Crude (%)	2337	17.01	0.0400	17.26	0.3595	0.1709	156	-0.70	002.05	0
002	Protein, Crude (%)	0038	17.02	0.5600	17.26	0.3595	0.1709	156	-0.67	002.06	0
002	Protein, Crude (%)	0881	17.02	0.0640	17.26	0.3595	0.1709	156	-0.66	002.05	0
002	Protein, Crude (%)	2259	17.03	0.0250	17.26	0.3595	0.1709	156	-0.66	002.01	0
002	Protein, Crude (%)	0968	17.04	0.1300	17.26	0.3595	0.1709	156	-0.63	002.06	0
002	Protein, Crude (%)	2340	17.04	0.1270	17.26	0.3595	0.1709	156	-0.62	002.05	0
002	Protein, Crude (%)	0100	17.06	0.0700	17.26	0.3595	0.1709	156	-0.58	002.06	0
002	Protein, Crude (%)	0870	17.06	0.1756	17.26	0.3595	0.1709	156	-0.56	002.06	0
002	Protein, Crude (%)	2109	17.07	0.1700	17.26	0.3595	0.1709	156	-0.55	002.06	0
002	Protein, Crude (%)	0354	17.08	0.0300	17.26	0.3595	0.1709	156	-0.52	002.05	0
002	Protein, Crude (%)	2081	17.08	0.2800	17.26	0.3595	0.1709	156	-0.51	002.06	0
002	Protein, Crude (%)	0027	17.09	0.0350	17.26	0.3595	0.1709	156	-0.47	002.06	0
002	Protein, Crude (%)	0848	17.10	0.2100	17.26	0.3595	0.1709	156	-0.46	002.06	0
002	Protein, Crude (%)	0968	17.10	0.0400	17.26	0.3595	0.1709	156	-0.45	002.01	0
002	Protein, Crude (%)	0510	17.10	0.0000	17.26	0.3595	0.1709	156	-0.45	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	2319	17.10	0.0000	17.26	0.3595	0.1709	156	-0.45	002.06	0
002	Protein, Crude (%)	0148	17.10	0.0110	17.26	0.3595	0.1709	156	-0.45	002.06	0
002	Protein, Crude (%)	0123	17.11	0.0600	17.26	0.3595	0.1709	156	-0.42	002.06	0
002	Protein, Crude (%)	2109	17.12	0.1400	17.26	0.3595	0.1709	156	-0.40	002.05	0
002	Protein, Crude (%)	0003	17.14	0.2600	17.26	0.3595	0.1709	156	-0.34	002.06	0
002	Protein, Crude (%)	2302	17.14	0.3400	17.26	0.3595	0.1709	156	-0.34	002.06	0
002	Protein, Crude (%)	0652	17.15	0.1000	17.26	0.3595	0.1709	156	-0.31	002.01	0
002	Protein, Crude (%)	0045	17.15	0.3000	17.26	0.3595	0.1709	156	-0.31	002.06	0
002	Protein, Crude (%)	0610	17.15	0.1000	17.26	0.3595	0.1709	156	-0.31	002.06	0
002	Protein, Crude (%)	0098	17.15	0.0400	17.26	0.3595	0.1709	156	-0.31	002.08	0
002	Protein, Crude (%)	2303	17.16	0.0900	17.26	0.3595	0.1709	156	-0.30	002.06	0
002	Protein, Crude (%)	2193	17.16	0.1000	17.26	0.3595	0.1709	156	-0.28	002.01	0
002	Protein, Crude (%)	2190	17.16	0.0800	17.26	0.3595	0.1709	156	-0.28	002.05	0
002	Protein, Crude (%)	2168	17.16	0.0200	17.26	0.3595	0.1709	156	-0.28	002.06	0
002	Protein, Crude (%)	0675	17.19	0.0500	17.26	0.3595	0.1709	156	-0.21	002.06	0
002	Protein, Crude (%)	0028	17.20	0.1700	17.26	0.3595	0.1709	156	-0.19	002.06	0
002	Protein, Crude (%)	0049	17.20	0.4000	17.26	0.3595	0.1709	156	-0.17	002.06	0
002	Protein, Crude (%)	0964	17.20	0.0000	17.26	0.3595	0.1709	156	-0.17	002.11	0
002	Protein, Crude (%)	0541	17.21	0.0600	17.26	0.3595	0.1709	156	-0.14	002.06	0
002	Protein, Crude (%)	0004	17.22	0.2500	17.26	0.3595	0.1709	156	-0.13	002.06	0
002	Protein, Crude (%)	0511	17.22	0.1400	17.26	0.3595	0.1709	156	-0.12	002.06	0
002	Protein, Crude (%)	0036	17.23	0.1110	17.26	0.3595	0.1709	156	-0.10	002.06	0
002	Protein, Crude (%)	0626	17.23	0.1460	17.26	0.3595	0.1709	156	-0.09	002.06	0
002	Protein, Crude (%)	2174	17.24	0.2200	17.26	0.3595	0.1709	156	-0.06	002.06	0
002	Protein, Crude (%)	0354	17.25	0.0100	17.26	0.3595	0.1709	156	-0.05	002.06	0
002	Protein, Crude (%)	0504	17.25	0.2100	17.26	0.3595	0.1709	156	-0.05	002.06	0
002	Protein, Crude (%)	0674	17.25	0.0000	17.26	0.3595	0.1709	156	-0.03	002.01	0
002	Protein, Crude (%)	0619	17.25	0.3000	17.26	0.3595	0.1709	156	-0.03	002.05	0
002	Protein, Crude (%)	0278	17.25	0.1000	17.26	0.3595	0.1709	156	-0.03	002.06	0
002	Protein, Crude (%)	0294	17.25	0.1000	17.26	0.3595	0.1709	156	-0.03	002.06	0
002	Protein, Crude (%)	0693	17.25	0.3750	17.26	0.3595	0.1709	156	-0.03	002.06	0
002	Protein, Crude (%)	0066	17.27	0.0900	17.26	0.3595	0.1709	156	0.01	002.02	0
002	Protein, Crude (%)	0171	17.27	0.0500	17.26	0.3595	0.1709	156	0.01	002.06	0
002	Protein, Crude (%)	2302	17.27	0.0000	17.26	0.3595	0.1709	156	0.02	002.11	0
002	Protein, Crude (%)	0019	17.28	0.0900	17.26	0.3595	0.1709	156	0.04	002.06	0
002	Protein, Crude (%)	0202	17.28	0.1300	17.26	0.3595	0.1709	156	0.04	002.06	0
002	Protein, Crude (%)	0662	17.28	0.1875	17.26	0.3595	0.1709	156	0.05	002.01	0
002	Protein, Crude (%)	0407	17.28	0.0010	17.26	0.3595	0.1709	156	0.06	002.06	0
002	Protein, Crude (%)	0358	17.29	0.1600	17.26	0.3595	0.1709	156	0.08	002.06	0
002	Protein, Crude (%)	0682	17.29	0.0000	17.26	0.3595	0.1709	156	0.08	002.06	0
002	Protein, Crude (%)	0508	17.30	0.4540	17.26	0.3595	0.1709	156	0.09	002.06	0
002	Protein, Crude (%)	0875	17.30	0.2000	17.26	0.3595	0.1709	156	0.11	002.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0366	17.30	0.2000	17.26	0.3595	0.1709	156	0.11	002.06	0
002	Protein, Crude (%)	0010	17.30	0.0000	17.26	0.3595	0.1709	156	0.11	002.06	0
002	Protein, Crude (%)	0017	17.30	0.0000	17.26	0.3595	0.1709	156	0.11	002.06	0
002	Protein, Crude (%)	2345	17.31	0.2200	17.26	0.3595	0.1709	156	0.13	002.00	0
002	Protein, Crude (%)	0589	17.32	0.0100	17.26	0.3595	0.1709	156	0.15	002.06	0
002	Protein, Crude (%)	0263	17.33	0.0820	17.26	0.3595	0.1709	156	0.18	002.06	0
002	Protein, Crude (%)	2059	17.33	0.1469	17.26	0.3595	0.1709	156	0.19	002.06	0
002	Protein, Crude (%)	0941	17.35	0.1100	17.26	0.3595	0.1709	156	0.23	002.05	0
002	Protein, Crude (%)	0910	17.35	0.5000	17.26	0.3595	0.1709	156	0.24	002.01	0
002	Protein, Crude (%)	0910	17.35	0.1000	17.26	0.3595	0.1709	156	0.24	002.06	0
002	Protein, Crude (%)	0964	17.35	0.0270	17.26	0.3595	0.1709	156	0.25	002.06	0
002	Protein, Crude (%)	0011	17.37	0.4700	17.26	0.3595	0.1709	156	0.29	002.06	0
002	Protein, Crude (%)	0726	17.37	0.1000	17.26	0.3595	0.1709	156	0.30	002.06	0
002	Protein, Crude (%)	0309	17.39	0.0960	17.26	0.3595	0.1709	156	0.36	002.06	0
002	Protein, Crude (%)	0098	17.40	0.4000	17.26	0.3595	0.1709	156	0.38	002.06	0
002	Protein, Crude (%)	0142	17.40	0.0200	17.26	0.3595	0.1709	156	0.38	002.06	0
002	Protein, Crude (%)	0687	17.40	0.2000	17.26	0.3595	0.1709	156	0.38	002.06	0
002	Protein, Crude (%)	0843	17.40	0.0200	17.26	0.3595	0.1709	156	0.38	002.06	0
002	Protein, Crude (%)	0035	17.41	0.1700	17.26	0.3595	0.1709	156	0.40	002.06	0
002	Protein, Crude (%)	0353	17.42	0.0100	17.26	0.3595	0.1709	156	0.43	002.06	0
002	Protein, Crude (%)	0563	17.42	0.5869	17.26	0.3595	0.1709	156	0.45	002.08	0
002	Protein, Crude (%)	0037	17.43	0.4300	17.26	0.3595	0.1709	156	0.45	002.06	0
002	Protein, Crude (%)	0651	17.44	0.4578	17.26	0.3595	0.1709	156	0.49	002.05	0
002	Protein, Crude (%)	0884	17.45	0.5000	17.26	0.3595	0.1709	156	0.52	002.06	0
002	Protein, Crude (%)	0574	17.46	0.0500	17.26	0.3595	0.1709	156	0.54	002.06	0
002	Protein, Crude (%)	0425	17.47	0.4400	17.26	0.3595	0.1709	156	0.58	002.06	0
002	Protein, Crude (%)	0089	17.48	0.0100	17.26	0.3595	0.1709	156	0.59	002.06	0
002	Protein, Crude (%)	0242	17.49	0.2100	17.26	0.3595	0.1709	156	0.62	002.06	0
002	Protein, Crude (%)	0619	17.50	0.2000	17.26	0.3595	0.1709	156	0.66	002.06	0
002	Protein, Crude (%)	2342	17.50	0.2000	17.26	0.3595	0.1709	156	0.66	002.99	0
002	Protein, Crude (%)	0108	17.53	0.0300	17.26	0.3595	0.1709	156	0.73	002.06	0
002	Protein, Crude (%)	0226	17.53	0.1300	17.26	0.3595	0.1709	156	0.73	002.06	0
002	Protein, Crude (%)	0914	17.54	0.0900	17.26	0.3595	0.1709	156	0.76	002.06	0
002	Protein, Crude (%)	0033	17.55	0.1000	17.26	0.3595	0.1709	156	0.80	002.06	0
002	Protein, Crude (%)	0956	17.55	0.1000	17.26	0.3595	0.1709	156	0.80	002.06	0
002	Protein, Crude (%)	0571	17.56	0.0100	17.26	0.3595	0.1709	156	0.81	002.06	0
002	Protein, Crude (%)	0001	17.57	0.0400	17.26	0.3595	0.1709	156	0.86	002.06	0
002	Protein, Crude (%)	2213	17.58	0.3200	17.26	0.3595	0.1709	156	0.88	002.06	0
002	Protein, Crude (%)	0553	17.59	0.3040	17.26	0.3595	0.1709	156	0.92	002.06	0
002	Protein, Crude (%)	0646	17.60	0.3700	17.26	0.3595	0.1709	156	0.93	002.06	0
002	Protein, Crude (%)	0208	17.60	0.2000	17.26	0.3595	0.1709	156	0.94	002.06	0
002	Protein, Crude (%)	2150	17.63	0.0400	17.26	0.3595	0.1709	156	1.02	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0169	17.64	0.4100	17.26	0.3595	0.1709	156	1.04	002.00	0
002	Protein, Crude (%)	0009	17.66	0.6214	17.26	0.3595	0.1709	156	1.11	002.06	0
002	Protein, Crude (%)	2089	17.70	0.0500	17.26	0.3595	0.1709	156	1.20	002.06	0
002	Protein, Crude (%)	0880	17.70	0.2000	17.26	0.3595	0.1709	156	1.22	002.06	0
002	Protein, Crude (%)	0190	17.73	0.0200	17.26	0.3595	0.1709	156	1.30	002.06	0
002	Protein, Crude (%)	0083	17.74	0.0700	17.26	0.3595	0.1709	156	1.32	002.06	0
002	Protein, Crude (%)	2076	17.75	0.3750	17.26	0.3595	0.1709	156	1.36	002.06	0
002	Protein, Crude (%)	0034	17.75	0.1160	17.26	0.3595	0.1709	156	1.37	002.06	0
002	Protein, Crude (%)	0018	17.78	0.2500	17.26	0.3595	0.1709	156	1.43	002.06	0
002	Protein, Crude (%)	0852	17.80	0.0100	17.26	0.3595	0.1709	156	1.48	002.11	0
002	Protein, Crude (%)	0692	18.00	0.0000	17.26	0.3595	0.1709	156	2.05	002.06	0
002	Protein, Crude (%)	0720	18.05	0.3000	17.26	0.3595	0.1709	156	2.19	002.06	0
002	Protein, Crude (%)	0345	18.05	0.1000	17.26	0.3595	0.1709	156	2.19	002.06	0
002	Protein, Crude (%)	2188	18.13	0.0000	17.26	0.3595	0.1709	156	2.41	002.01	0
002	Protein, Crude (%)	0720	18.17	0.3000	17.26	0.3595	0.1709	156	2.53	002.11	0
002	Protein, Crude (%)	0300	18.25	0.3000	17.26	0.3595	0.1709	156	2.75	002.06	0
002	Protein, Crude (%)	0265	18.30	0.2000	17.26	0.3595	0.1709	156	2.89	002.06	0
002	Protein, Crude (%)	2155	18.33	0.0990	17.26	0.3595	0.1709	156	2.97	002.06	0
002	Protein, Crude (%)	2326	18.35	0.0700	17.26	0.3595	0.1709	156	3.01	002.06	0
002	Protein, Crude (%)	0852	18.39	0.0000	17.26	0.3595	0.1709	156	3.14	002.05	0
002	Protein, Crude (%)	0581	18.40	0.2000	17.26	0.3595	0.1709	156	3.16	002.05	0
002	Protein, Crude (%)	2178	18.56	0.7100	17.26	0.3595	0.1709	156	3.60	002.06	0
002	Protein, Crude (%)	0011	18.92	0.0000	17.26	0.3595	0.1709	156	4.61	002.11	0
002	Protein, Crude (%)	0612	22.00	0.6200	17.26	0.3595	0.1709	156	13.18	002.06	0
002	Protein, Crude (%)	0859	15.55	2.849	17.26	0.3595	0.1709	156	-4.77	002.06	1
002	Protein, Crude (%)	0042	19.60	1.200	17.26	0.3595	0.1709	156	6.50	002.06	1
002	Protein, Crude (%)	2129	19.66	1.400	17.26	0.3595	0.1709	156	6.67	002.06	1
002	Protein, Crude (%)	0297	34.84	0.0800	17.26	0.3595	0.1709	156	48.89	002.06	2
002	Protein, Crude (%)	0853	35.80	0.8000	17.26	0.3595	0.1709	156	51.56	002.06	2
002	Protein, Crude (%)	0297	41.82	0.1200	17.26	0.3595	0.1709	156	68.30	002.11	2
003	Fat, Crude (%)	0011	3.155	0.2900	3.857	0.2681	0.1013	125	-2.62	003.00	0
003	Fat, Crude (%)	0357	3.305	0.0100	3.857	0.2681	0.1013	125	-2.06	003.14	0
003	Fat, Crude (%)	0619	3.325	0.1100	3.857	0.2681	0.1013	125	-1.99	003.10	0
003	Fat, Crude (%)	0011	3.330	0.0000	3.857	0.2681	0.1013	125	-1.97	003.11	0
003	Fat, Crude (%)	2073	3.395	0.0500	3.857	0.2681	0.1013	125	-1.72	003.10	0
003	Fat, Crude (%)	0693	3.404	0.0230	3.857	0.2681	0.1013	125	-1.69	003.10	0
003	Fat, Crude (%)	0345	3.415	0.1900	3.857	0.2681	0.1013	125	-1.65	003.00	0
003	Fat, Crude (%)	0042	3.425	0.2300	3.857	0.2681	0.1013	125	-1.61	003.10	0
003	Fat, Crude (%)	0100	3.460	0.2000	3.857	0.2681	0.1013	125	-1.48	003.10	0
003	Fat, Crude (%)	0876	3.500	0.2000	3.857	0.2681	0.1013	125	-1.33	003.14	0
003	Fat, Crude (%)	0504	3.525	0.2300	3.857	0.2681	0.1013	125	-1.24	003.01	0
003	Fat, Crude (%)	0529	3.540	0.0400	3.857	0.2681	0.1013	125	-1.18	003.14	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
003	Fat, Crude (%)	0366	3.550	0.1000	3.857	0.2681	0.1013	125	-1.15	003.10	0
003	Fat, Crude (%)	0956	3.550	0.1000	3.857	0.2681	0.1013	125	-1.15	003.14	0
003	Fat, Crude (%)	2193	3.560	0.0200	3.857	0.2681	0.1013	125	-1.11	003.10	0
003	Fat, Crude (%)	2089	3.570	0.0200	3.857	0.2681	0.1013	125	-1.07	003.10	0
003	Fat, Crude (%)	0574	3.575	0.1100	3.857	0.2681	0.1013	125	-1.05	003.14	0
003	Fat, Crude (%)	0644	3.576	0.0960	3.857	0.2681	0.1013	125	-1.05	003.06	0
003	Fat, Crude (%)	2337	3.590	0.1400	3.857	0.2681	0.1013	125	-1.00	003.00	0
003	Fat, Crude (%)	0263	3.597	0.0700	3.857	0.2681	0.1013	125	-0.97	003.10	0
003	Fat, Crude (%)	2009	3.598	0.0268	3.857	0.2681	0.1013	125	-0.97	003.10	0
003	Fat, Crude (%)	0689	3.600	0.0000	3.857	0.2681	0.1013	125	-0.96	003.10	0
003	Fat, Crude (%)	2196	3.610	0.0200	3.857	0.2681	0.1013	125	-0.92	003.10	0
003	Fat, Crude (%)	0036	3.614	0.0205	3.857	0.2681	0.1013	125	-0.91	003.10	0
003	Fat, Crude (%)	0939	3.620	0.1800	3.857	0.2681	0.1013	125	-0.89	003.14	0
003	Fat, Crude (%)	2184	3.625	0.0300	3.857	0.2681	0.1013	125	-0.87	003.10	0
003	Fat, Crude (%)	0089	3.625	0.0700	3.857	0.2681	0.1013	125	-0.87	003.14	0
003	Fat, Crude (%)	0941	3.625	0.1300	3.857	0.2681	0.1013	125	-0.87	003.99	0
003	Fat, Crude (%)	0662	3.640	0.1558	3.857	0.2681	0.1013	125	-0.81	003.10	0
003	Fat, Crude (%)	0682	3.640	0.0000	3.857	0.2681	0.1013	125	-0.81	003.14	0
003	Fat, Crude (%)	0019	3.645	0.0700	3.857	0.2681	0.1013	125	-0.79	003.14	0
003	Fat, Crude (%)	0964	3.650	0.1000	3.857	0.2681	0.1013	125	-0.77	003.11	0
003	Fat, Crude (%)	0689	3.700	0.0000	3.857	0.2681	0.1013	125	-0.59	003.06	0
003	Fat, Crude (%)	0034	3.710	0.1200	3.857	0.2681	0.1013	125	-0.55	003.14	0
003	Fat, Crude (%)	0037	3.715	0.0300	3.857	0.2681	0.1013	125	-0.53	003.14	0
003	Fat, Crude (%)	0300	3.740	0.0600	3.857	0.2681	0.1013	125	-0.44	003.14	0
003	Fat, Crude (%)	0358	3.740	0.0200	3.857	0.2681	0.1013	125	-0.44	003.14	0
003	Fat, Crude (%)	0242	3.750	0.0200	3.857	0.2681	0.1013	125	-0.40	003.10	0
003	Fat, Crude (%)	0265	3.750	0.2800	3.857	0.2681	0.1013	125	-0.40	003.14	0
003	Fat, Crude (%)	0581	3.750	0.1000	3.857	0.2681	0.1013	125	-0.40	003.14	0
003	Fat, Crude (%)	0066	3.755	0.0500	3.857	0.2681	0.1013	125	-0.38	003.14	0
003	Fat, Crude (%)	0870	3.755	0.0858	3.857	0.2681	0.1013	125	-0.38	003.10	0
003	Fat, Crude (%)	2188	3.760	0.0000	3.857	0.2681	0.1013	125	-0.36	003.10	0
003	Fat, Crude (%)	0511	3.760	0.0600	3.857	0.2681	0.1013	125	-0.36	003.14	0
003	Fat, Crude (%)	0098	3.765	0.1100	3.857	0.2681	0.1013	125	-0.34	003.10	0
003	Fat, Crude (%)	0202	3.780	0.1800	3.857	0.2681	0.1013	125	-0.29	003.14	0
003	Fat, Crude (%)	2340	3.790	0.0220	3.857	0.2681	0.1013	125	-0.25	003.00	0
003	Fat, Crude (%)	0108	3.790	0.0200	3.857	0.2681	0.1013	125	-0.25	003.14	0
003	Fat, Crude (%)	2155	3.799	0.0850	3.857	0.2681	0.1013	125	-0.22	003.10	0
003	Fat, Crude (%)	0045	3.800	0.2000	3.857	0.2681	0.1013	125	-0.21	003.10	0
003	Fat, Crude (%)	0510	3.800	0.2800	3.857	0.2681	0.1013	125	-0.21	003.14	0
003	Fat, Crude (%)	2076	3.805	0.3563	3.857	0.2681	0.1013	125	-0.20	003.14	0
003	Fat, Crude (%)	0407	3.808	0.1733	3.857	0.2681	0.1013	125	-0.18	003.14	0
003	Fat, Crude (%)	0033	3.810	0.1000	3.857	0.2681	0.1013	125	-0.18	003.13	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
003	Fat, Crude (%)	0148	3.815	0.0300	3.857	0.2681	0.1013	125	-0.16	003.10	0
003	Fat, Crude (%)	0294	3.820	0.0600	3.857	0.2681	0.1013	125	-0.14	003.06	0
003	Fat, Crude (%)	0425	3.825	0.1100	3.857	0.2681	0.1013	125	-0.12	003.06	0
003	Fat, Crude (%)	0541	3.830	0.2000	3.857	0.2681	0.1013	125	-0.10	003.10	0
003	Fat, Crude (%)	0881	3.837	0.0880	3.857	0.2681	0.1013	125	-0.08	003.10	0
003	Fat, Crude (%)	2109	3.850	0.1400	3.857	0.2681	0.1013	125	-0.03	003.00	0
003	Fat, Crude (%)	0884	3.850	0.1000	3.857	0.2681	0.1013	125	-0.03	003.06	0
003	Fat, Crude (%)	0142	3.855	0.0900	3.857	0.2681	0.1013	125	-0.01	003.00	0
003	Fat, Crude (%)	0948	3.855	0.0100	3.857	0.2681	0.1013	125	-0.01	003.09	0
003	Fat, Crude (%)	0148	3.855	0.0700	3.857	0.2681	0.1013	125	-0.01	003.14	0
003	Fat, Crude (%)	0171	3.855	0.2700	3.857	0.2681	0.1013	125	-0.01	003.14	0
003	Fat, Crude (%)	0171	3.860	0.0400	3.857	0.2681	0.1013	125	0.01	003.12	0
003	Fat, Crude (%)	0536	3.860	0.0600	3.857	0.2681	0.1013	125	0.01	003.99	0
003	Fat, Crude (%)	0169	3.875	0.1100	3.857	0.2681	0.1013	125	0.07	003.06	0
003	Fat, Crude (%)	2302	3.880	0.0000	3.857	0.2681	0.1013	125	0.08	003.11	0
003	Fat, Crude (%)	0144	3.880	0.0600	3.857	0.2681	0.1013	125	0.08	003.14	0
003	Fat, Crude (%)	0948	3.880	0.0800	3.857	0.2681	0.1013	125	0.08	003.14	0
003	Fat, Crude (%)	0619	3.885	0.0500	3.857	0.2681	0.1013	125	0.10	003.14	0
003	Fat, Crude (%)	0910	3.890	0.1800	3.857	0.2681	0.1013	125	0.12	003.06	0
003	Fat, Crude (%)	0208	3.895	0.1300	3.857	0.2681	0.1013	125	0.14	003.13	0
003	Fat, Crude (%)	0035	3.895	0.0300	3.857	0.2681	0.1013	125	0.14	003.14	0
003	Fat, Crude (%)	0083	3.895	0.0100	3.857	0.2681	0.1013	125	0.14	003.14	0
003	Fat, Crude (%)	2319	3.900	0.0000	3.857	0.2681	0.1013	125	0.16	003.10	0
003	Fat, Crude (%)	0001	3.906	0.1485	3.857	0.2681	0.1013	125	0.18	003.14	0
003	Fat, Crude (%)	0848	3.910	0.0800	3.857	0.2681	0.1013	125	0.20	003.14	0
003	Fat, Crude (%)	0505	3.920	0.1800	3.857	0.2681	0.1013	125	0.23	003.14	0
003	Fat, Crude (%)	0004	3.925	0.2100	3.857	0.2681	0.1013	125	0.25	003.09	0
003	Fat, Crude (%)	0028	3.925	0.2500	3.857	0.2681	0.1013	125	0.25	003.13	0
003	Fat, Crude (%)	0123	3.940	0.1200	3.857	0.2681	0.1013	125	0.31	003.13	0
003	Fat, Crude (%)	0009	3.940	0.0400	3.857	0.2681	0.1013	125	0.31	003.14	0
003	Fat, Crude (%)	2006	3.955	0.1100	3.857	0.2681	0.1013	125	0.36	003.13	0
003	Fat, Crude (%)	2302	3.955	0.1100	3.857	0.2681	0.1013	125	0.36	003.14	0
003	Fat, Crude (%)	0598	3.970	0.0600	3.857	0.2681	0.1013	125	0.42	003.14	0
003	Fat, Crude (%)	0098	3.970	0.0400	3.857	0.2681	0.1013	125	0.42	003.13	0
003	Fat, Crude (%)	2326	3.975	0.0100	3.857	0.2681	0.1013	125	0.44	003.12	0
003	Fat, Crude (%)	0651	3.976	0.5167	3.857	0.2681	0.1013	125	0.44	003.14	0
003	Fat, Crude (%)	0674	3.980	0.0000	3.857	0.2681	0.1013	125	0.46	003.06	0
003	Fat, Crude (%)	0004	3.995	0.0300	3.857	0.2681	0.1013	125	0.51	003.14	0
003	Fat, Crude (%)	0692	4.000	0.2000	3.857	0.2681	0.1013	125	0.53	003.14	0
003	Fat, Crude (%)	0027	4.008	0.0320	3.857	0.2681	0.1013	125	0.56	003.09	0
003	Fat, Crude (%)	2325	4.035	0.0500	3.857	0.2681	0.1013	125	0.66	003.00	0
003	Fat, Crude (%)	0675	4.045	0.0700	3.857	0.2681	0.1013	125	0.70	003.14	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
003	Fat, Crude (%)	0720	4.050	0.0600	3.857	0.2681	0.1013	125	0.72	003.11	0
003	Fat, Crude (%)	0297	4.050	0.0400	3.857	0.2681	0.1013	125	0.72	003.06	0
003	Fat, Crude (%)	0098	4.070	0.1000	3.857	0.2681	0.1013	125	0.79	003.09	0
003	Fat, Crude (%)	0687	4.075	0.1500	3.857	0.2681	0.1013	125	0.81	003.06	0
003	Fat, Crude (%)	0646	4.075	0.0700	3.857	0.2681	0.1013	125	0.81	003.13	0
003	Fat, Crude (%)	0354	4.090	0.0000	3.857	0.2681	0.1013	125	0.87	003.09	0
003	Fat, Crude (%)	0309	4.095	0.0215	3.857	0.2681	0.1013	125	0.89	003.00	0
003	Fat, Crude (%)	0563	4.114	0.0780	3.857	0.2681	0.1013	125	0.96	003.01	0
003	Fat, Crude (%)	0563	4.114	0.0780	3.857	0.2681	0.1013	125	0.96	003.14	0
003	Fat, Crude (%)	2192	4.135	0.3500	3.857	0.2681	0.1013	125	1.04	003.00	0
003	Fat, Crude (%)	0968	4.135	0.0100	3.857	0.2681	0.1013	125	1.04	003.14	0
003	Fat, Crude (%)	0853	4.150	0.1000	3.857	0.2681	0.1013	125	1.09	003.14	0
003	Fat, Crude (%)	0682	4.170	0.0000	3.857	0.2681	0.1013	125	1.17	003.06	0
003	Fat, Crude (%)	0726	4.235	0.0100	3.857	0.2681	0.1013	125	1.41	003.14	0
003	Fat, Crude (%)	0278	4.250	0.1000	3.857	0.2681	0.1013	125	1.46	003.14	0
003	Fat, Crude (%)	0035	4.275	0.1100	3.857	0.2681	0.1013	125	1.56	003.00	0
003	Fat, Crude (%)	2190	4.295	0.1300	3.857	0.2681	0.1013	125	1.63	003.00	0
003	Fat, Crude (%)	0017	4.300	0.0000	3.857	0.2681	0.1013	125	1.65	003.00	0
003	Fat, Crude (%)	2342	4.300	0.2000	3.857	0.2681	0.1013	125	1.65	003.09	0
003	Fat, Crude (%)	0885	4.389	0.1181	3.857	0.2681	0.1013	125	1.98	003.00	0
003	Fat, Crude (%)	0870	4.423	0.0445	3.857	0.2681	0.1013	125	2.11	003.13	0
003	Fat, Crude (%)	0852	4.450	0.1000	3.857	0.2681	0.1013	125	2.21	003.11	0
003	Fat, Crude (%)	2129	4.450	0.6800	3.857	0.2681	0.1013	125	2.21	003.14	0
003	Fat, Crude (%)	0353	4.470	0.0200	3.857	0.2681	0.1013	125	2.28	003.09	0
003	Fat, Crude (%)	0190	4.500	0.0400	3.857	0.2681	0.1013	125	2.40	003.14	0
003	Fat, Crude (%)	0226	4.570	0.0600	3.857	0.2681	0.1013	125	2.66	003.14	0
003	Fat, Crude (%)	0297	4.685	0.1300	3.857	0.2681	0.1013	125	3.09	003.11	0
003	Fat, Crude (%)	0520	4.710	0.0400	3.857	0.2681	0.1013	125	3.18	003.14	0
003	Fat, Crude (%)	0843	6.040	0.2600	3.857	0.2681	0.1013	125	8.14	003.14	0
003	Fat, Crude (%)	0049	2.735	0.7700	3.857	0.2681	0.1013	125	-4.19	003.14	1
003	Fat, Crude (%)	0003	3.740	0.7400	3.857	0.2681	0.1013	125	-0.44	003.14	1
003	Fat, Crude (%)	0074	4.370	0.8200	3.857	0.2681	0.1013	125	1.91	003.14	1
003	Fat, Crude (%)	0038	6.120	1.000	3.857	0.2681	0.1013	125	8.44	003.09	1
003	Fat, Crude (%)	0880	0.5000	1.000	3.857	0.2681	0.1013	125		003.99	4
004	Fiber, Crude (%)	0852	12.88	0.1700	20.34	0.9153	0.3952	106	-8.16	004.11	0
004	Fiber, Crude (%)	0720	16.45	0.3000	20.34	0.9153	0.3952	106	-4.25	004.11	0
004	Fiber, Crude (%)	0407	17.43	0.0088	20.34	0.9153	0.3952	106	-3.18	004.07	0
004	Fiber, Crude (%)	0619	17.75	0.1000	20.34	0.9153	0.3952	106	-2.83	004.07	0
004	Fiber, Crude (%)	0353	18.35	0.3000	20.34	0.9153	0.3952	106	-2.18	004.03	0
004	Fiber, Crude (%)	2006	18.85	0.2800	20.34	0.9153	0.3952	106	-1.63	004.06	0
004	Fiber, Crude (%)	0011	18.92	0.0000	20.34	0.9153	0.3952	106	-1.55	004.11	0
004	Fiber, Crude (%)	0956	18.95	0.5000	20.34	0.9153	0.3952	106	-1.52	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
004	Fiber, Crude (%)	0644	19.17	1.087	20.34	0.9153	0.3952	106	-1.28	004.07	0
004	Fiber, Crude (%)	0726	19.20	0.0300	20.34	0.9153	0.3952	106	-1.25	004.07	0
004	Fiber, Crude (%)	0968	19.27	0.1100	20.34	0.9153	0.3952	106	-1.18	004.06	0
004	Fiber, Crude (%)	0169	19.35	0.1000	20.34	0.9153	0.3952	106	-1.08	004.00	0
004	Fiber, Crude (%)	0202	19.35	0.3000	20.34	0.9153	0.3952	106	-1.08	004.07	0
004	Fiber, Crude (%)	2326	19.42	0.0700	20.34	0.9153	0.3952	106	-1.01	004.00	0
004	Fiber, Crude (%)	2319	19.45	0.1000	20.34	0.9153	0.3952	106	-0.97	004.06	0
004	Fiber, Crude (%)	0881	19.47	0.0590	20.34	0.9153	0.3952	106	-0.96	004.06	0
004	Fiber, Crude (%)	0885	19.48	0.0233	20.34	0.9153	0.3952	106	-0.94	004.06	0
004	Fiber, Crude (%)	0881	19.48	0.0600	20.34	0.9153	0.3952	106	-0.94	004.07	0
004	Fiber, Crude (%)	0100	19.51	0.2400	20.34	0.9153	0.3952	106	-0.91	004.07	0
004	Fiber, Crude (%)	0144	19.52	0.0800	20.34	0.9153	0.3952	106	-0.90	004.07	0
004	Fiber, Crude (%)	0884	19.60	0.4000	20.34	0.9153	0.3952	106	-0.81	004.00	0
004	Fiber, Crude (%)	0689	19.65	0.3000	20.34	0.9153	0.3952	106	-0.76	004.06	0
004	Fiber, Crude (%)	0278	19.65	0.1000	20.34	0.9153	0.3952	106	-0.76	004.07	0
004	Fiber, Crude (%)	0226	19.66	0.4000	20.34	0.9153	0.3952	106	-0.75	004.07	0
004	Fiber, Crude (%)	2089	19.69	0.2100	20.34	0.9153	0.3952	106	-0.72	004.03	0
004	Fiber, Crude (%)	0028	19.70	0.8000	20.34	0.9153	0.3952	106	-0.70	004.07	0
004	Fiber, Crude (%)	2342	19.75	0.7000	20.34	0.9153	0.3952	106	-0.65	004.07	0
004	Fiber, Crude (%)	0964	19.75	0.5000	20.34	0.9153	0.3952	106	-0.65	004.11	0
004	Fiber, Crude (%)	0098	19.83	1.010	20.34	0.9153	0.3952	106	-0.56	004.06	0
004	Fiber, Crude (%)	0171	19.84	0.3800	20.34	0.9153	0.3952	106	-0.55	004.00	0
004	Fiber, Crude (%)	0581	19.85	0.1000	20.34	0.9153	0.3952	106	-0.54	004.07	0
004	Fiber, Crude (%)	2340	19.88	0.0260	20.34	0.9153	0.3952	106	-0.50	004.07	0
004	Fiber, Crude (%)	0968	19.90	0.3100	20.34	0.9153	0.3952	106	-0.49	004.07	0
004	Fiber, Crude (%)	0682	19.90	0.0000	20.34	0.9153	0.3952	106	-0.48	004.07	0
004	Fiber, Crude (%)	0294	19.95	0.1000	20.34	0.9153	0.3952	106	-0.43	004.07	0
004	Fiber, Crude (%)	0505	20.02	0.1200	20.34	0.9153	0.3952	106	-0.35	004.07	0
004	Fiber, Crude (%)	0043	20.04	0.1100	20.34	0.9153	0.3952	106	-0.34	004.07	0
004	Fiber, Crude (%)	0190	20.05	0.3000	20.34	0.9153	0.3952	106	-0.32	004.07	0
004	Fiber, Crude (%)	0425	20.05	0.1000	20.34	0.9153	0.3952	106	-0.32	004.00	0
004	Fiber, Crude (%)	0300	20.05	0.9000	20.34	0.9153	0.3952	106	-0.32	004.07	0
004	Fiber, Crude (%)	0910	20.05	0.5000	20.34	0.9153	0.3952	106	-0.32	004.07	0
004	Fiber, Crude (%)	0027	20.07	0.6420	20.34	0.9153	0.3952	106	-0.30	004.06	0
004	Fiber, Crude (%)	0563	20.08	0.0982	20.34	0.9153	0.3952	106	-0.29	004.07	0
004	Fiber, Crude (%)	2302	20.09	0.2100	20.34	0.9153	0.3952	106	-0.28	004.07	0
004	Fiber, Crude (%)	0035	20.13	0.1200	20.34	0.9153	0.3952	106	-0.23	004.07	0
004	Fiber, Crude (%)	0643	20.14	0.3550	20.34	0.9153	0.3952	106	-0.22	004.07	0
004	Fiber, Crude (%)	0610	20.15	0.1000	20.34	0.9153	0.3952	106	-0.21	004.06	0
004	Fiber, Crude (%)	2325	20.16	0.0300	20.34	0.9153	0.3952	106	-0.20	004.06	0
004	Fiber, Crude (%)	0004	20.18	0.4800	20.34	0.9153	0.3952	106	-0.18	004.07	0
004	Fiber, Crude (%)	0876	20.20	0.4000	20.34	0.9153	0.3952	106	-0.16	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
004	Fiber, Crude (%)	0345	20.20	0.8000	20.34	0.9153	0.3952	106	-0.16	004.00	0
004	Fiber, Crude (%)	0309	20.25	0.2687	20.34	0.9153	0.3952	106	-0.10	004.00	0
004	Fiber, Crude (%)	0510	20.25	0.1000	20.34	0.9153	0.3952	106	-0.10	004.07	0
004	Fiber, Crude (%)	0674	20.27	0.0000	20.34	0.9153	0.3952	106	-0.08	004.06	0
004	Fiber, Crude (%)	0083	20.27	0.1200	20.34	0.9153	0.3952	106	-0.08	004.07	0
004	Fiber, Crude (%)	0263	20.29	0.0280	20.34	0.9153	0.3952	106	-0.06	004.07	0
004	Fiber, Crude (%)	2259	20.29	1.269	20.34	0.9153	0.3952	106	-0.05	004.07	0
004	Fiber, Crude (%)	0074	20.31	0.8100	20.34	0.9153	0.3952	106	-0.04	004.07	0
004	Fiber, Crude (%)	2337	20.35	0.0500	20.34	0.9153	0.3952	106	0.00	004.07	0
004	Fiber, Crude (%)	0693	20.35	1.455	20.34	0.9153	0.3952	106	0.01	004.07	0
004	Fiber, Crude (%)	0939	20.37	0.2600	20.34	0.9153	0.3952	106	0.03	004.07	0
004	Fiber, Crude (%)	0692	20.45	0.3000	20.34	0.9153	0.3952	106	0.12	004.07	0
004	Fiber, Crude (%)	0646	20.46	1.880	20.34	0.9153	0.3952	106	0.13	004.07	0
004	Fiber, Crude (%)	0038	20.53	0.8500	20.34	0.9153	0.3952	106	0.20	004.07	0
004	Fiber, Crude (%)	2192	20.54	0.5300	20.34	0.9153	0.3952	106	0.21	004.03	0
004	Fiber, Crude (%)	0009	20.56	0.3000	20.34	0.9153	0.3952	106	0.24	004.07	0
004	Fiber, Crude (%)	0848	20.57	0.5000	20.34	0.9153	0.3952	106	0.25	004.07	0
004	Fiber, Crude (%)	0017	20.60	0.4000	20.34	0.9153	0.3952	106	0.28	004.00	0
004	Fiber, Crude (%)	0208	20.60	0.6000	20.34	0.9153	0.3952	106	0.28	004.00	0
004	Fiber, Crude (%)	0529	20.62	0.1000	20.34	0.9153	0.3952	106	0.30	004.07	0
004	Fiber, Crude (%)	0676	20.69	0.6300	20.34	0.9153	0.3952	106	0.37	004.06	0
004	Fiber, Crude (%)	0948	20.79	0.1800	20.34	0.9153	0.3952	106	0.49	004.06	0
004	Fiber, Crude (%)	0511	20.80	1.200	20.34	0.9153	0.3952	106	0.50	004.07	0
004	Fiber, Crude (%)	0589	20.82	0.2800	20.34	0.9153	0.3952	106	0.52	004.07	0
004	Fiber, Crude (%)	0626	20.85	0.4800	20.34	0.9153	0.3952	106	0.56	004.07	0
004	Fiber, Crude (%)	0520	20.87	0.1200	20.34	0.9153	0.3952	106	0.58	004.07	0
004	Fiber, Crude (%)	0045	20.90	0.0000	20.34	0.9153	0.3952	106	0.61	004.07	0
004	Fiber, Crude (%)	0066	20.91	0.4500	20.34	0.9153	0.3952	106	0.62	004.07	0
004	Fiber, Crude (%)	2193	20.91	0.0300	20.34	0.9153	0.3952	106	0.62	004.07	0
004	Fiber, Crude (%)	0870	20.93	0.4628	20.34	0.9153	0.3952	106	0.64	004.07	0
004	Fiber, Crude (%)	2302	20.94	0.0000	20.34	0.9153	0.3952	106	0.65	004.11	0
004	Fiber, Crude (%)	0098	20.95	1.920	20.34	0.9153	0.3952	106	0.66	004.07	0
004	Fiber, Crude (%)	0504	20.96	0.8100	20.34	0.9153	0.3952	106	0.67	004.00	0
004	Fiber, Crude (%)	0366	21.00	1.200	20.34	0.9153	0.3952	106	0.72	004.07	0
004	Fiber, Crude (%)	0675	21.04	0.1900	20.34	0.9153	0.3952	106	0.76	004.07	0
004	Fiber, Crude (%)	2196	21.10	0.0000	20.34	0.9153	0.3952	106	0.83	004.07	0
004	Fiber, Crude (%)	0019	21.17	0.1800	20.34	0.9153	0.3952	106	0.90	004.07	0
004	Fiber, Crude (%)	0553	21.24	0.7500	20.34	0.9153	0.3952	106	0.98	004.07	0
004	Fiber, Crude (%)	0123	21.28	0.1100	20.34	0.9153	0.3952	106	1.02	004.07	0
004	Fiber, Crude (%)	0089	21.37	0.0400	20.34	0.9153	0.3952	106	1.12	004.07	0
004	Fiber, Crude (%)	0941	21.40	0.2000	20.34	0.9153	0.3952	106	1.16	004.06	0
004	Fiber, Crude (%)	0011	21.40	0.8000	20.34	0.9153	0.3952	106	1.16	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
004	Fiber, Crude (%)	0536	21.49	0.1100	20.34	0.9153	0.3952	106	1.25	004.07	0
004	Fiber, Crude (%)	0034	21.52	0.0000	20.34	0.9153	0.3952	106	1.29	004.07	0
004	Fiber, Crude (%)	2073	21.57	0.1200	20.34	0.9153	0.3952	106	1.34	004.00	0
004	Fiber, Crude (%)	0003	21.73	1.390	20.34	0.9153	0.3952	106	1.51	004.07	0
004	Fiber, Crude (%)	0598	21.76	0.3500	20.34	0.9153	0.3952	106	1.54	004.07	0
004	Fiber, Crude (%)	0720	21.80	0.6900	20.34	0.9153	0.3952	106	1.59	004.06	0
004	Fiber, Crude (%)	0914	21.90	0.2000	20.34	0.9153	0.3952	106	1.70	004.07	0
004	Fiber, Crude (%)	0265	21.95	1.700	20.34	0.9153	0.3952	106	1.76	004.07	0
004	Fiber, Crude (%)	0354	22.01	0.2900	20.34	0.9153	0.3952	106	1.82	004.07	0
004	Fiber, Crude (%)	2109	22.31	0.1400	20.34	0.9153	0.3952	106	2.15	004.07	0
004	Fiber, Crude (%)	2129	22.94	1.315	20.34	0.9153	0.3952	106	2.84	004.99	0
004	Fiber, Crude (%)	2345	22.96	1.190	20.34	0.9153	0.3952	106	2.85	004.07	0
004	Fiber, Crude (%)	2076	23.63	0.0190	20.34	0.9153	0.3952	106	3.60	004.00	0
004	Fiber, Crude (%)	2188	26.04	0.0000	20.34	0.9153	0.3952	106	6.23	004.07	0
004	Fiber, Crude (%)	0042	22.11	1.980	20.34	0.9153	0.3952	106	1.93	004.07	1
005	Ash (%)	0852	8.925	0.1500	11.18	0.1906	0.0855	102	-11.85	005.00	0
005	Ash (%)	0108	9.480	0.3000	11.18	0.1906	0.0855	102	-8.94	005.00	0
005	Ash (%)	0852	9.670	0.2400	11.18	0.1906	0.0855	102	-7.95	005.11	0
005	Ash (%)	2081	10.07	0.2600	11.18	0.1906	0.0855	102	-5.85	005.99	0
005	Ash (%)	2345	10.55	0.1000	11.18	0.1906	0.0855	102	-3.33	005.05	0
005	Ash (%)	0644	10.67	0.0460	11.18	0.1906	0.0855	102	-2.69	005.02	0
005	Ash (%)	0510	10.69	0.0700	11.18	0.1906	0.0855	102	-2.62	005.00	0
005	Ash (%)	0169	10.75	0.0100	11.18	0.1906	0.0855	102	-2.31	005.00	0
005	Ash (%)	0853	10.80	0.4000	11.18	0.1906	0.0855	102	-2.02	005.00	0
005	Ash (%)	0045	10.85	0.3000	11.18	0.1906	0.0855	102	-1.76	005.00	0
005	Ash (%)	0278	10.85	0.0800	11.18	0.1906	0.0855	102	-1.76	005.00	0
005	Ash (%)	0358	10.93	0.1100	11.18	0.1906	0.0855	102	-1.36	005.00	0
005	Ash (%)	0357	10.95	0.1000	11.18	0.1906	0.0855	102	-1.23	005.00	0
005	Ash (%)	0425	10.97	0.1500	11.18	0.1906	0.0855	102	-1.15	005.00	0
005	Ash (%)	0144	10.98	0.0400	11.18	0.1906	0.0855	102	-1.07	005.00	0
005	Ash (%)	0589	10.99	0.0300	11.18	0.1906	0.0855	102	-1.05	005.00	0
005	Ash (%)	0208	11.00	0.2000	11.18	0.1906	0.0855	102	-0.97	005.00	0
005	Ash (%)	0651	11.01	0.0175	11.18	0.1906	0.0855	102	-0.92	005.00	0
005	Ash (%)	2184	11.02	0.1200	11.18	0.1906	0.0855	102	-0.86	005.00	0
005	Ash (%)	0142	11.04	0.0300	11.18	0.1906	0.0855	102	-0.79	005.00	0
005	Ash (%)	0049	11.04	0.0400	11.18	0.1906	0.0855	102	-0.76	005.00	0
005	Ash (%)	0019	11.05	0.1000	11.18	0.1906	0.0855	102	-0.71	005.00	0
005	Ash (%)	2188	11.05	0.0000	11.18	0.1906	0.0855	102	-0.71	005.00	0
005	Ash (%)	2268	11.06	0.0300	11.18	0.1906	0.0855	102	-0.68	005.00	0
005	Ash (%)	0407	11.06	0.0980	11.18	0.1906	0.0855	102	-0.66	005.00	0
005	Ash (%)	0171	11.06	0.0200	11.18	0.1906	0.0855	102	-0.65	005.00	0
005	Ash (%)	0953	11.06	0.3000	11.18	0.1906	0.0855	102	-0.65	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
005	Ash (%)	0042	11.09	0.0700	11.18	0.1906	0.0855	102	-0.52	005.00	0
005	Ash (%)	0726	11.09	0.0000	11.18	0.1906	0.0855	102	-0.50	005.05	0
005	Ash (%)	0539	11.10	0.0500	11.18	0.1906	0.0855	102	-0.47	005.00	0
005	Ash (%)	2193	11.10	0.0100	11.18	0.1906	0.0855	102	-0.47	005.00	0
005	Ash (%)	2340	11.10	0.0460	11.18	0.1906	0.0855	102	-0.47	005.00	0
005	Ash (%)	0956	11.10	0.2000	11.18	0.1906	0.0855	102	-0.44	005.00	0
005	Ash (%)	0366	11.10	0.4000	11.18	0.1906	0.0855	102	-0.44	005.00	0
005	Ash (%)	2302	11.12	0.0500	11.18	0.1906	0.0855	102	-0.37	005.00	0
005	Ash (%)	2006	11.12	0.0600	11.18	0.1906	0.0855	102	-0.34	005.00	0
005	Ash (%)	2076	11.13	0.0291	11.18	0.1906	0.0855	102	-0.29	005.00	0
005	Ash (%)	0598	11.13	0.0200	11.18	0.1906	0.0855	102	-0.29	005.00	0
005	Ash (%)	0035	11.14	0.0300	11.18	0.1906	0.0855	102	-0.26	005.00	0
005	Ash (%)	0574	11.15	0.1600	11.18	0.1906	0.0855	102	-0.18	005.00	0
005	Ash (%)	0881	11.15	0.1090	11.18	0.1906	0.0855	102	-0.18	005.00	0
005	Ash (%)	2337	11.16	0.0900	11.18	0.1906	0.0855	102	-0.16	005.00	0
005	Ash (%)	0968	11.16	0.0100	11.18	0.1906	0.0855	102	-0.16	005.00	0
005	Ash (%)	0720	11.16	0.0400	11.18	0.1906	0.0855	102	-0.13	005.05	0
005	Ash (%)	0643	11.17	0.0900	11.18	0.1906	0.0855	102	-0.10	005.00	0
005	Ash (%)	0689	11.17	0.0100	11.18	0.1906	0.0855	102	-0.10	005.05	0
005	Ash (%)	0870	11.17	0.1511	11.18	0.1906	0.0855	102	-0.10	005.00	0
005	Ash (%)	0083	11.18	0.0200	11.18	0.1906	0.0855	102	-0.02	005.00	0
005	Ash (%)	0541	11.18	0.0600	11.18	0.1906	0.0855	102	-0.02	005.00	0
005	Ash (%)	0682	11.18	0.0000	11.18	0.1906	0.0855	102	-0.02	005.00	0
005	Ash (%)	2196	11.18	0.0000	11.18	0.1906	0.0855	102	-0.02	005.00	0
005	Ash (%)	0646	11.19	0.0800	11.18	0.1906	0.0855	102	0.03	005.00	0
005	Ash (%)	2150	11.19	0.1000	11.18	0.1906	0.0855	102	0.03	005.00	0
005	Ash (%)	0034	11.19	0.0413	11.18	0.1906	0.0855	102	0.04	005.05	0
005	Ash (%)	0294	11.20	0.0000	11.18	0.1906	0.0855	102	0.08	005.05	0
005	Ash (%)	0226	11.21	0.0700	11.18	0.1906	0.0855	102	0.11	005.00	0
005	Ash (%)	0948	11.21	0.0100	11.18	0.1906	0.0855	102	0.11	005.00	0
005	Ash (%)	0520	11.21	0.0200	11.18	0.1906	0.0855	102	0.13	005.00	0
005	Ash (%)	2073	11.21	0.0200	11.18	0.1906	0.0855	102	0.13	005.00	0
005	Ash (%)	0309	11.22	0.2600	11.18	0.1906	0.0855	102	0.19	005.00	0
005	Ash (%)	0242	11.23	0.0700	11.18	0.1906	0.0855	102	0.21	005.00	0
005	Ash (%)	2326	11.23	0.1300	11.18	0.1906	0.0855	102	0.21	005.00	0
005	Ash (%)	0354	11.23	0.0100	11.18	0.1906	0.0855	102	0.21	005.00	0
005	Ash (%)	0693	11.23	0.1260	11.18	0.1906	0.0855	102	0.24	005.00	0
005	Ash (%)	0263	11.23	0.0020	11.18	0.1906	0.0855	102	0.25	005.05	0
005	Ash (%)	0536	11.24	0.0100	11.18	0.1906	0.0855	102	0.26	005.99	0
005	Ash (%)	0563	11.26	0.1075	11.18	0.1906	0.0855	102	0.37	005.00	0
005	Ash (%)	2259	11.26	0.0431	11.18	0.1906	0.0855	102	0.39	005.00	0
005	Ash (%)	0504	11.27	0.0700	11.18	0.1906	0.0855	102	0.42	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
005	Ash (%)	0880	11.27	0.0400	11.18	0.1906	0.0855	102	0.45	005.00	0
005	Ash (%)	0100	11.30	0.0000	11.18	0.1906	0.0855	102	0.60	005.00	0
005	Ash (%)	0848	11.30	0.0400	11.18	0.1906	0.0855	102	0.60	005.00	0
005	Ash (%)	2089	11.30	0.0200	11.18	0.1906	0.0855	102	0.60	005.00	0
005	Ash (%)	0610	11.30	0.0000	11.18	0.1906	0.0855	102	0.60	005.05	0
005	Ash (%)	0939	11.31	0.0300	11.18	0.1906	0.0855	102	0.63	005.05	0
005	Ash (%)	0553	11.31	0.0600	11.18	0.1906	0.0855	102	0.66	005.00	0
005	Ash (%)	0148	11.32	0.1000	11.18	0.1906	0.0855	102	0.71	005.00	0
005	Ash (%)	0674	11.32	0.0000	11.18	0.1906	0.0855	102	0.71	005.00	0
005	Ash (%)	0123	11.34	0.0900	11.18	0.1906	0.0855	102	0.79	005.00	0
005	Ash (%)	0644	11.35	0.0540	11.18	0.1906	0.0855	102	0.84	005.05	0
005	Ash (%)	0676	11.35	0.0000	11.18	0.1906	0.0855	102	0.87	005.99	0
005	Ash (%)	0265	11.35	0.1000	11.18	0.1906	0.0855	102	0.87	005.00	0
005	Ash (%)	2342	11.35	0.1000	11.18	0.1906	0.0855	102	0.87	005.00	0
005	Ash (%)	0910	11.35	0.1000	11.18	0.1906	0.0855	102	0.87	005.99	0
005	Ash (%)	0098	11.36	0.0500	11.18	0.1906	0.0855	102	0.89	005.00	0
005	Ash (%)	2109	11.36	0.1800	11.18	0.1906	0.0855	102	0.92	005.05	0
005	Ash (%)	0941	11.38	0.1700	11.18	0.1906	0.0855	102	1.00	005.05	0
005	Ash (%)	2190	11.39	0.0700	11.18	0.1906	0.0855	102	1.05	005.05	0
005	Ash (%)	0692	11.40	0.2000	11.18	0.1906	0.0855	102	1.13	005.05	0
005	Ash (%)	0202	11.40	0.0000	11.18	0.1906	0.0855	102	1.13	005.99	0
005	Ash (%)	0884	11.40	0.2000	11.18	0.1906	0.0855	102	1.13	005.99	0
005	Ash (%)	2009	11.42	0.0231	11.18	0.1906	0.0855	102	1.26	005.05	0
005	Ash (%)	0353	11.43	0.0700	11.18	0.1906	0.0855	102	1.26	005.00	0
005	Ash (%)	0914	11.43	0.0300	11.18	0.1906	0.0855	102	1.26	005.00	0
005	Ash (%)	0885	11.43	0.0058	11.18	0.1906	0.0855	102	1.28	005.05	0
005	Ash (%)	0652	11.45	0.1000	11.18	0.1906	0.0855	102	1.39	005.99	0
005	Ash (%)	0345	11.51	0.2100	11.18	0.1906	0.0855	102	1.68	005.00	0
005	Ash (%)	2325	11.53	0.0300	11.18	0.1906	0.0855	102	1.79	005.00	0
005	Ash (%)	0675	11.53	0.0200	11.18	0.1906	0.0855	102	1.81	005.00	0
005	Ash (%)	0581	11.55	0.1000	11.18	0.1906	0.0855	102	1.92	005.00	0
005	Ash (%)	2129	12.71	0.1020	11.18	0.1906	0.0855	102	7.99	005.05	0
005	Ash (%)	0297	21.56	0.2100	11.18	0.1906	0.0855	102	54.40	005.00	0
005	Ash (%)	0720	8.505	0.6100	11.18	0.1906	0.0855	102	-14.06	005.11	1
005	Ash (%)	0297	26.83	0.3000	11.18	0.1906	0.0855	102	82.07	005.11	2
006	Total Sugars (%)	0720	1.970	0.0400	3.800	1.097	0.2080	5		006.99	0
006	Total Sugars (%)	0407	3.630	0.3200	3.800	1.097	0.2080	5		006.00	0
006	Total Sugars (%)	2192	4.215	0.2700	3.800	1.097	0.2080	5		006.00	0
006	Total Sugars (%)	0226	4.535	0.1100	3.800	1.097	0.2080	5		006.99	0
006	Total Sugars (%)	0956	4.650	0.3000	3.800	1.097	0.2080	5		006.99	0
008	Fiber, Acid Detergent (%)	2302	21.94	0.2100	27.17	1.777	0.3889	53	-2.95	008.08	0
008	Fiber, Acid Detergent (%)	0910	23.92	0.0100	27.17	1.777	0.3889	53	-1.83	008.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
008	Fiber, Acid Detergent (%)	0675	23.93	0.2400	27.17	1.777	0.3889	53	-1.82	008.08	0
008	Fiber, Acid Detergent (%)	0956	23.95	0.1000	27.17	1.777	0.3889	53	-1.81	008.08	0
008	Fiber, Acid Detergent (%)	0610	24.10	0.4000	27.17	1.777	0.3889	53	-1.73	008.99	0
008	Fiber, Acid Detergent (%)	0354	24.33	0.0200	27.17	1.777	0.3889	53	-1.60	008.08	0
008	Fiber, Acid Detergent (%)	0407	24.42	0.4081	27.17	1.777	0.3889	53	-1.55	008.08	0
008	Fiber, Acid Detergent (%)	0948	24.65	0.6700	27.17	1.777	0.3889	53	-1.42	008.02	0
008	Fiber, Acid Detergent (%)	0674	25.03	0.0000	27.17	1.777	0.3889	53	-1.20	008.02	0
008	Fiber, Acid Detergent (%)	0968	25.04	0.3600	27.17	1.777	0.3889	53	-1.20	008.08	0
008	Fiber, Acid Detergent (%)	0941	25.10	0.2600	27.17	1.777	0.3889	53	-1.16	008.99	0
008	Fiber, Acid Detergent (%)	0848	25.15	0.4700	27.17	1.777	0.3889	53	-1.14	008.08	0
008	Fiber, Acid Detergent (%)	0042	25.87	0.1300	27.17	1.777	0.3889	53	-0.73	008.08	0
008	Fiber, Acid Detergent (%)	0581	25.90	0.4000	27.17	1.777	0.3889	53	-0.71	008.08	0
008	Fiber, Acid Detergent (%)	2196	26.27	0.0000	27.17	1.777	0.3889	53	-0.51	008.08	0
008	Fiber, Acid Detergent (%)	0720	26.57	0.0000	27.17	1.777	0.3889	53	-0.34	008.99	0
008	Fiber, Acid Detergent (%)	0689	26.75	0.7000	27.17	1.777	0.3889	53	-0.24	008.02	0
008	Fiber, Acid Detergent (%)	0049	27.02	0.1600	27.17	1.777	0.3889	53	-0.08	008.08	0
008	Fiber, Acid Detergent (%)	0001	27.18	0.0150	27.17	1.777	0.3889	53	0.01	008.08	0
008	Fiber, Acid Detergent (%)	0004	27.21	2.070	27.17	1.777	0.3889	53	0.02	008.08	0
008	Fiber, Acid Detergent (%)	0190	27.25	0.0000	27.17	1.777	0.3889	53	0.05	008.08	0
008	Fiber, Acid Detergent (%)	0357	27.33	0.1500	27.17	1.777	0.3889	53	0.09	008.08	0
008	Fiber, Acid Detergent (%)	0066	27.44	0.5300	27.17	1.777	0.3889	53	0.15	008.08	0
008	Fiber, Acid Detergent (%)	0510	27.45	0.1000	27.17	1.777	0.3889	53	0.16	008.08	0
008	Fiber, Acid Detergent (%)	0098	27.46	0.3300	27.17	1.777	0.3889	53	0.16	008.02	0
008	Fiber, Acid Detergent (%)	0646	27.53	0.3300	27.17	1.777	0.3889	53	0.20	008.08	0
008	Fiber, Acid Detergent (%)	0278	27.55	0.1000	27.17	1.777	0.3889	53	0.21	008.08	0
008	Fiber, Acid Detergent (%)	0294	27.55	0.1000	27.17	1.777	0.3889	53	0.21	008.08	0
008	Fiber, Acid Detergent (%)	0263	27.63	0.0250	27.17	1.777	0.3889	53	0.26	008.08	0
008	Fiber, Acid Detergent (%)	0345	27.75	0.3000	27.17	1.777	0.3889	53	0.33	008.02	0
008	Fiber, Acid Detergent (%)	2129	27.80	1.800	27.17	1.777	0.3889	53	0.35	008.08	0
008	Fiber, Acid Detergent (%)	0083	27.82	0.3100	27.17	1.777	0.3889	53	0.36	008.08	0
008	Fiber, Acid Detergent (%)	0309	27.85	0.4349	27.17	1.777	0.3889	53	0.38	008.02	0
008	Fiber, Acid Detergent (%)	0676	27.88	0.7400	27.17	1.777	0.3889	53	0.40	008.02	0
008	Fiber, Acid Detergent (%)	0870	27.88	1.626	27.17	1.777	0.3889	53	0.40	008.08	0
008	Fiber, Acid Detergent (%)	0964	28.00	0.2000	27.17	1.777	0.3889	53	0.47	008.99	0
008	Fiber, Acid Detergent (%)	0226	28.12	0.6900	27.17	1.777	0.3889	53	0.53	008.02	0
008	Fiber, Acid Detergent (%)	0353	28.20	0.3800	27.17	1.777	0.3889	53	0.58	008.02	0
008	Fiber, Acid Detergent (%)	0511	28.25	0.1000	27.17	1.777	0.3889	53	0.61	008.08	0
008	Fiber, Acid Detergent (%)	0035	28.39	0.1800	27.17	1.777	0.3889	53	0.69	008.08	0
008	Fiber, Acid Detergent (%)	0358	28.47	0.1900	27.17	1.777	0.3889	53	0.73	008.08	0
008	Fiber, Acid Detergent (%)	0536	28.47	0.0000	27.17	1.777	0.3889	53	0.73	008.08	0
008	Fiber, Acid Detergent (%)	0148	28.55	0.1300	27.17	1.777	0.3889	53	0.77	008.02	0
008	Fiber, Acid Detergent (%)	0037	28.67	0.4000	27.17	1.777	0.3889	53	0.84	008.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
008	Fiber, Acid Detergent (%)	0693	28.72	0.5540	27.17	1.777	0.3889	53	0.87	008.08	0
008	Fiber, Acid Detergent (%)	0504	28.73	0.7600	27.17	1.777	0.3889	53	0.88	008.02	0
008	Fiber, Acid Detergent (%)	0726	28.84	1.420	27.17	1.777	0.3889	53	0.94	008.08	0
008	Fiber, Acid Detergent (%)	0265	28.85	0.1000	27.17	1.777	0.3889	53	0.95	008.05	0
008	Fiber, Acid Detergent (%)	0148	28.85	0.9800	27.17	1.777	0.3889	53	0.95	008.08	0
008	Fiber, Acid Detergent (%)	0964	29.02	0.0480	27.17	1.777	0.3889	53	1.04	008.02	0
008	Fiber, Acid Detergent (%)	2193	29.30	0.1400	27.17	1.777	0.3889	53	1.20	008.08	0
008	Fiber, Acid Detergent (%)	0045	29.35	0.7000	27.17	1.777	0.3889	53	1.23	008.08	0
008	Fiber, Acid Detergent (%)	2326	31.62	0.1400	27.17	1.777	0.3889	53	2.50	008.99	0
008	Fiber, Acid Detergent (%)	0939	26.24	2.400	27.17	1.777	0.3889	53	-0.52	008.08	1
008	Fiber, Acid Detergent (%)	0297	10.37	1.020	27.17	1.777	0.3889	53	-9.45	008.08	2
009	Fiber, Neutral Detergent (%)	0407	33.33	0.0420	36.59	1.624	0.3743	50	-2.01	009.09	0
009	Fiber, Neutral Detergent (%)	0263	33.70	0.6420	36.59	1.624	0.3743	50	-1.77	009.09	0
009	Fiber, Neutral Detergent (%)	2302	34.05	0.5000	36.59	1.624	0.3743	50	-1.56	009.09	0
009	Fiber, Neutral Detergent (%)	0294	34.25	0.3000	36.59	1.624	0.3743	50	-1.44	009.09	0
009	Fiber, Neutral Detergent (%)	0598	34.43	0.1300	36.59	1.624	0.3743	50	-1.33	009.09	0
009	Fiber, Neutral Detergent (%)	0910	34.70	0.2000	36.59	1.624	0.3743	50	-1.16	009.07	0
009	Fiber, Neutral Detergent (%)	2196	34.93	0.0000	36.59	1.624	0.3743	50	-1.02	009.09	0
009	Fiber, Neutral Detergent (%)	0357	34.95	0.1000	36.59	1.624	0.3743	50	-1.01	009.09	0
009	Fiber, Neutral Detergent (%)	0646	35.08	0.0500	36.59	1.624	0.3743	50	-0.93	009.09	0
009	Fiber, Neutral Detergent (%)	0581	35.15	0.3000	36.59	1.624	0.3743	50	-0.88	009.09	0
009	Fiber, Neutral Detergent (%)	0358	35.41	0.0200	36.59	1.624	0.3743	50	-0.72	009.09	0
009	Fiber, Neutral Detergent (%)	0354	35.51	0.0200	36.59	1.624	0.3743	50	-0.66	009.09	0
009	Fiber, Neutral Detergent (%)	0035	35.53	0.2300	36.59	1.624	0.3743	50	-0.65	009.09	0
009	Fiber, Neutral Detergent (%)	0536	35.56	0.0300	36.59	1.624	0.3743	50	-0.63	009.09	0
009	Fiber, Neutral Detergent (%)	2193	35.62	0.1400	36.59	1.624	0.3743	50	-0.59	009.09	0
009	Fiber, Neutral Detergent (%)	0726	35.75	0.7600	36.59	1.624	0.3743	50	-0.51	009.09	0
009	Fiber, Neutral Detergent (%)	2340	35.92	0.0030	36.59	1.624	0.3743	50	-0.41	009.09	0
009	Fiber, Neutral Detergent (%)	0675	35.95	0.3900	36.59	1.624	0.3743	50	-0.39	009.09	0
009	Fiber, Neutral Detergent (%)	0720	35.95	0.0500	36.59	1.624	0.3743	50	-0.39	009.99	0
009	Fiber, Neutral Detergent (%)	0510	36.00	0.2000	36.59	1.624	0.3743	50	-0.36	009.09	0
009	Fiber, Neutral Detergent (%)	0353	36.17	0.2300	36.59	1.624	0.3743	50	-0.26	009.07	0
009	Fiber, Neutral Detergent (%)	0956	36.20	0.6000	36.59	1.624	0.3743	50	-0.24	009.09	0
009	Fiber, Neutral Detergent (%)	0939	36.30	0.2000	36.59	1.624	0.3743	50	-0.18	009.09	0
009	Fiber, Neutral Detergent (%)	0083	36.37	0.2100	36.59	1.624	0.3743	50	-0.14	009.09	0
009	Fiber, Neutral Detergent (%)	2326	36.47	0.1000	36.59	1.624	0.3743	50	-0.07	009.99	0
009	Fiber, Neutral Detergent (%)	0870	36.48	0.8086	36.59	1.624	0.3743	50	-0.07	009.09	0
009	Fiber, Neutral Detergent (%)	0693	36.74	0.7980	36.59	1.624	0.3743	50	0.09	009.09	0
009	Fiber, Neutral Detergent (%)	0345	36.96	0.1400	36.59	1.624	0.3743	50	0.23	009.07	0
009	Fiber, Neutral Detergent (%)	0948	36.99	0.0600	36.59	1.624	0.3743	50	0.25	009.07	0
009	Fiber, Neutral Detergent (%)	2129	37.10	0.4000	36.59	1.624	0.3743	50	0.32	009.09	0
009	Fiber, Neutral Detergent (%)	0037	37.14	0.8500	36.59	1.624	0.3743	50	0.34	009.09	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
009	Fiber, Neutral Detergent (%)	0045	37.15	0.5000	36.59	1.624	0.3743	50	0.35	009.09	0
009	Fiber, Neutral Detergent (%)	0066	37.20	0.4900	36.59	1.624	0.3743	50	0.38	009.09	0
009	Fiber, Neutral Detergent (%)	0674	37.20	0.0000	36.59	1.624	0.3743	50	0.38	009.09	0
009	Fiber, Neutral Detergent (%)	0148	37.35	0.0400	36.59	1.624	0.3743	50	0.47	009.09	0
009	Fiber, Neutral Detergent (%)	0504	37.38	0.1200	36.59	1.624	0.3743	50	0.49	009.04	0
009	Fiber, Neutral Detergent (%)	0610	37.50	0.6000	36.59	1.624	0.3743	50	0.56	009.07	0
009	Fiber, Neutral Detergent (%)	0309	37.68	0.2053	36.59	1.624	0.3743	50	0.67	009.07	0
009	Fiber, Neutral Detergent (%)	0278	37.85	0.1000	36.59	1.624	0.3743	50	0.78	009.09	0
009	Fiber, Neutral Detergent (%)	0968	37.94	0.4000	36.59	1.624	0.3743	50	0.83	009.09	0
009	Fiber, Neutral Detergent (%)	0226	38.03	0.1200	36.59	1.624	0.3743	50	0.89	009.07	0
009	Fiber, Neutral Detergent (%)	0689	38.15	1.100	36.59	1.624	0.3743	50	0.96	009.07	0
009	Fiber, Neutral Detergent (%)	0964	38.22	0.4740	36.59	1.624	0.3743	50	1.00	009.07	0
009	Fiber, Neutral Detergent (%)	0098	38.30	1.200	36.59	1.624	0.3743	50	1.06	009.07	0
009	Fiber, Neutral Detergent (%)	0941	38.49	0.2700	36.59	1.624	0.3743	50	1.17	009.99	0
009	Fiber, Neutral Detergent (%)	0265	38.75	1.700	36.59	1.624	0.3743	50	1.33	009.09	0
009	Fiber, Neutral Detergent (%)	0190	39.48	0.5600	36.59	1.624	0.3743	50	1.78	009.09	0
009	Fiber, Neutral Detergent (%)	0848	40.79	0.1500	36.59	1.624	0.3743	50	2.59	009.09	0
009	Fiber, Neutral Detergent (%)	0964	41.20	1.000	36.59	1.624	0.3743	50	2.84	009.99	0
009	Fiber, Neutral Detergent (%)	0676	41.64	1.180	36.59	1.624	0.3743	50	3.11	009.07	0
009	Fiber, Neutral Detergent (%)	0049	37.31	2.340	36.59	1.624	0.3743	50	0.45	009.09	1
009	Fiber, Neutral Detergent (%)	0297	18.12	1.040	36.59	1.624	0.3743	50	-11.37	009.09	2
010	Moisture (%)	0941	6.555	0.1900	8.506	0.5519	0.1001	21	-3.54	010.99	0
010	Moisture (%)	0964	7.400	0.2000	8.506	0.5519	0.1001	21	-2.00	010.11	0
010	Moisture (%)	0692	7.910	0.2200	8.506	0.5519	0.1001	21	-1.08	010.99	0
010	Moisture (%)	2043	8.060	0.0000	8.506	0.5519	0.1001	21	-0.81	010.99	0
010	Moisture (%)	2089	8.125	0.3100	8.506	0.5519	0.1001	21	-0.69	010.99	0
010	Moisture (%)	0948	8.235	0.0300	8.506	0.5519	0.1001	21	-0.49	010.99	0
010	Moisture (%)	0939	8.380	0.0000	8.506	0.5519	0.1001	21	-0.23	010.99	0
010	Moisture (%)	0726	8.390	0.0000	8.506	0.5519	0.1001	21	-0.21	010.99	0
010	Moisture (%)	2150	8.420	0.0400	8.506	0.5519	0.1001	21	-0.16	010.99	0
010	Moisture (%)	2178	8.435	0.3100	8.506	0.5519	0.1001	21	-0.13	010.99	0
010	Moisture (%)	0921	8.435	0.0180	8.506	0.5519	0.1001	21	-0.13	010.99	0
010	Moisture (%)	2076	8.505	0.1300	8.506	0.5519	0.1001	21	0.00	010.99	0
010	Moisture (%)	0875	8.530	0.0000	8.506	0.5519	0.1001	21	0.04	010.99	0
010	Moisture (%)	0884	8.550	0.1000	8.506	0.5519	0.1001	21	0.08	010.99	0
010	Moisture (%)	0652	8.600	0.0000	8.506	0.5519	0.1001	21	0.17	010.99	0
010	Moisture (%)	0190	8.795	0.0300	8.506	0.5519	0.1001	21	0.52	010.99	0
010	Moisture (%)	0720	8.920	0.0800	8.506	0.5519	0.1001	21	0.75	010.11	0
010	Moisture (%)	0914	8.988	0.2950	8.506	0.5519	0.1001	21	0.87	010.99	0
010	Moisture (%)	2342	9.330	0.0600	8.506	0.5519	0.1001	21	1.49	010.99	0
010	Moisture (%)	2302	10.40	0.0000	8.506	0.5519	0.1001	21	3.43	010.11	0
010	Moisture (%)	0852	11.58	0.0900	8.506	0.5519	0.1001	21	5.56	010.11	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
010	Moisture (%)	2129	7.700	1.200	8.506	0.5519	0.1001	21	-1.46	010.99	1
010	Moisture (%)	0843	7.820	1.160	8.506	0.5519	0.1001	21	-1.24	010.03	1
011	Loss on Drying, HT (%)	0859	7.117	0.1640	9.407	0.4485	0.1109	65	-5.11	011.01	0
011	Loss on Drying, HT (%)	2322	8.210	0.0200	9.407	0.4485	0.1109	65	-2.67	011.01	0
011	Loss on Drying, HT (%)	2325	8.480	0.0000	9.407	0.4485	0.1109	65	-2.07	011.01	0
011	Loss on Drying, HT (%)	0047	8.700	0.0000	9.407	0.4485	0.1109	65	-1.58	011.01	0
011	Loss on Drying, HT (%)	0563	8.745	0.0881	9.407	0.4485	0.1109	65	-1.48	011.01	0
011	Loss on Drying, HT (%)	0953	8.775	0.1700	9.407	0.4485	0.1109	65	-1.41	011.01	0
011	Loss on Drying, HT (%)	0520	8.795	0.0500	9.407	0.4485	0.1109	65	-1.37	011.01	0
011	Loss on Drying, HT (%)	0536	8.845	0.0500	9.407	0.4485	0.1109	65	-1.25	011.01	0
011	Loss on Drying, HT (%)	0511	8.945	1.430	9.407	0.4485	0.1109	65	-1.03	011.01	0
011	Loss on Drying, HT (%)	2340	8.990	0.0580	9.407	0.4485	0.1109	65	-0.93	011.01	0
011	Loss on Drying, HT (%)	0265	9.000	0.0000	9.407	0.4485	0.1109	65	-0.91	011.01	0
011	Loss on Drying, HT (%)	0354	9.005	0.0100	9.407	0.4485	0.1109	65	-0.90	011.01	0
011	Loss on Drying, HT (%)	0208	9.030	0.1800	9.407	0.4485	0.1109	65	-0.84	011.01	0
011	Loss on Drying, HT (%)	2326	9.030	0.0200	9.407	0.4485	0.1109	65	-0.84	011.01	0
011	Loss on Drying, HT (%)	0942	9.030	0.0600	9.407	0.4485	0.1109	65	-0.84	011.02	0
011	Loss on Drying, HT (%)	0263	9.061	0.0050	9.407	0.4485	0.1109	65	-0.77	011.01	0
011	Loss on Drying, HT (%)	0598	9.095	0.0100	9.407	0.4485	0.1109	65	-0.70	011.01	0
011	Loss on Drying, HT (%)	2174	9.120	0.1600	9.407	0.4485	0.1109	65	-0.64	011.01	0
011	Loss on Drying, HT (%)	0968	9.120	0.0400	9.407	0.4485	0.1109	65	-0.64	011.01	0
011	Loss on Drying, HT (%)	2168	9.120	0.1400	9.407	0.4485	0.1109	65	-0.64	011.01	0
011	Loss on Drying, HT (%)	0100	9.130	0.1200	9.407	0.4485	0.1109	65	-0.62	011.01	0
011	Loss on Drying, HT (%)	2337	9.170	0.0200	9.407	0.4485	0.1109	65	-0.53	011.01	0
011	Loss on Drying, HT (%)	0539	9.260	0.2000	9.407	0.4485	0.1109	65	-0.33	011.01	0
011	Loss on Drying, HT (%)	0674	9.280	0.0000	9.407	0.4485	0.1109	65	-0.28	011.01	0
011	Loss on Drying, HT (%)	2006	9.315	0.0100	9.407	0.4485	0.1109	65	-0.21	011.01	0
011	Loss on Drying, HT (%)	2174	9.325	0.0100	9.407	0.4485	0.1109	65	-0.18	011.99	0
011	Loss on Drying, HT (%)	0425	9.335	0.0100	9.407	0.4485	0.1109	65	-0.16	011.01	0
011	Loss on Drying, HT (%)	0309	9.345	0.2100	9.407	0.4485	0.1109	65	-0.14	011.01	0
011	Loss on Drying, HT (%)	2193	9.360	0.0200	9.407	0.4485	0.1109	65	-0.11	011.01	0
011	Loss on Drying, HT (%)	2196	9.360	0.0000	9.407	0.4485	0.1109	65	-0.11	011.01	0
011	Loss on Drying, HT (%)	0651	9.363	0.2406	9.407	0.4485	0.1109	65	-0.10	011.01	0
011	Loss on Drying, HT (%)	0881	9.398	0.1420	9.407	0.4485	0.1109	65	-0.02	011.01	0
011	Loss on Drying, HT (%)	0226	9.410	0.1400	9.407	0.4485	0.1109	65	0.01	011.01	0
011	Loss on Drying, HT (%)	0294	9.430	0.0400	9.407	0.4485	0.1109	65	0.05	011.01	0
011	Loss on Drying, HT (%)	2073	9.465	0.0300	9.407	0.4485	0.1109	65	0.13	011.01	0
011	Loss on Drying, HT (%)	0123	9.480	0.1200	9.407	0.4485	0.1109	65	0.16	011.01	0
011	Loss on Drying, HT (%)	0066	9.485	0.2100	9.407	0.4485	0.1109	65	0.17	011.01	0
011	Loss on Drying, HT (%)	0646	9.515	0.1500	9.407	0.4485	0.1109	65	0.24	011.01	0
011	Loss on Drying, HT (%)	0870	9.516	0.0912	9.407	0.4485	0.1109	65	0.24	011.01	0
011	Loss on Drying, HT (%)	0643	9.545	0.0300	9.407	0.4485	0.1109	65	0.31	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
011	Loss on Drying, HT (%)	0541	9.560	0.1800	9.407	0.4485	0.1109	65	0.34	011.01	0
011	Loss on Drying, HT (%)	0407	9.565	0.0885	9.407	0.4485	0.1109	65	0.35	011.01	0
011	Loss on Drying, HT (%)	0675	9.640	0.0200	9.407	0.4485	0.1109	65	0.52	011.01	0
011	Loss on Drying, HT (%)	0529	9.650	0.0400	9.407	0.4485	0.1109	65	0.54	011.02	0
011	Loss on Drying, HT (%)	0574	9.660	0.1200	9.407	0.4485	0.1109	65	0.56	011.01	0
011	Loss on Drying, HT (%)	0553	9.675	0.2300	9.407	0.4485	0.1109	65	0.60	011.01	0
011	Loss on Drying, HT (%)	2268	9.675	0.2900	9.407	0.4485	0.1109	65	0.60	011.01	0
011	Loss on Drying, HT (%)	2302	9.675	0.0700	9.407	0.4485	0.1109	65	0.60	011.01	0
011	Loss on Drying, HT (%)	0852	9.695	0.3300	9.407	0.4485	0.1109	65	0.64	011.99	0
011	Loss on Drying, HT (%)	0171	9.705	0.0100	9.407	0.4485	0.1109	65	0.66	011.01	0
011	Loss on Drying, HT (%)	0843	9.710	0.0800	9.407	0.4485	0.1109	65	0.67	011.01	0
011	Loss on Drying, HT (%)	2168	9.750	0.0800	9.407	0.4485	0.1109	65	0.76	011.99	0
011	Loss on Drying, HT (%)	0682	9.840	0.0000	9.407	0.4485	0.1109	65	0.96	011.01	0
011	Loss on Drying, HT (%)	2303	9.841	0.2300	9.407	0.4485	0.1109	65	0.97	011.99	0
011	Loss on Drying, HT (%)	0098	9.845	0.2900	9.407	0.4485	0.1109	65	0.98	011.01	0
011	Loss on Drying, HT (%)	0242	9.880	0.0600	9.407	0.4485	0.1109	65	1.05	011.01	0
011	Loss on Drying, HT (%)	0144	9.890	0.2000	9.407	0.4485	0.1109	65	1.08	011.01	0
011	Loss on Drying, HT (%)	2184	9.895	0.0100	9.407	0.4485	0.1109	65	1.09	011.01	0
011	Loss on Drying, HT (%)	0589	9.945	0.1900	9.407	0.4485	0.1109	65	1.20	011.01	0
011	Loss on Drying, HT (%)	0510	9.950	0.1000	9.407	0.4485	0.1109	65	1.21	011.01	0
011	Loss on Drying, HT (%)	2188	10.03	0.0000	9.407	0.4485	0.1109	65	1.39	011.01	0
011	Loss on Drying, HT (%)	0004	10.04	0.0500	9.407	0.4485	0.1109	65	1.40	011.01	0
011	Loss on Drying, HT (%)	0358	10.08	0.0200	9.407	0.4485	0.1109	65	1.50	011.01	0
011	Loss on Drying, HT (%)	0202	10.20	0.0000	9.407	0.4485	0.1109	65	1.77	011.01	0
011	Loss on Drying, HT (%)	0108	11.32	0.0700	9.407	0.4485	0.1109	65	4.25	011.01	0
011	Loss on Drying, HT (%)	0880	9.515	3.690	9.407	0.4485	0.1109	65	0.24	011.01	1
012	Starch (%)	0297	2.560	0.1400	8.876	1.311	0.2403	25	-4.82	012.11	0
012	Starch (%)	0297	3.640	0.1800	8.876	1.311	0.2403	25	-3.99	012.03	0
012	Starch (%)	0407	7.311	0.1274	8.876	1.311	0.2403	25	-1.19	012.03	0
012	Starch (%)	0510	7.500	0.4000	8.876	1.311	0.2403	25	-1.05	012.04	0
012	Starch (%)	0870	7.670	0.1148	8.876	1.311	0.2403	25	-0.92	012.01	0
012	Starch (%)	0956	7.950	0.3000	8.876	1.311	0.2403	25	-0.71	012.01	0
012	Starch (%)	0278	8.000	0.2000	8.876	1.311	0.2403	25	-0.67	012.04	0
012	Starch (%)	0148	8.335	0.0900	8.876	1.311	0.2403	25	-0.41	012.01	0
012	Starch (%)	0045	8.380	0.2800	8.876	1.311	0.2403	25	-0.38	012.01	0
012	Starch (%)	0226	8.480	0.2000	8.876	1.311	0.2403	25	-0.30	012.04	0
012	Starch (%)	0693	8.619	0.8670	8.876	1.311	0.2403	25	-0.20	012.01	0
012	Starch (%)	2129	8.650	0.3200	8.876	1.311	0.2403	25	-0.17	012.04	0
012	Starch (%)	0353	8.870	0.0600	8.876	1.311	0.2403	25	0.00	012.20	0
012	Starch (%)	2192	8.900	0.1000	8.876	1.311	0.2403	25	0.02	012.00	0
012	Starch (%)	0098	8.990	1.500	8.876	1.311	0.2403	25	0.09	012.04	0
012	Starch (%)	0689	9.350	0.3000	8.876	1.311	0.2403	25	0.36	012.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
012	Starch (%)	2193	9.455	0.0100	8.876	1.311	0.2403	25	0.44	012.00	0
012	Starch (%)	0265	9.735	0.2100	8.876	1.311	0.2403	25	0.66	012.01	0
012	Starch (%)	2006	9.795	0.0100	8.876	1.311	0.2403	25	0.70	012.00	0
012	Starch (%)	2188	9.935	0.0300	8.876	1.311	0.2403	25	0.81	012.00	0
012	Starch (%)	0354	9.965	0.0500	8.876	1.311	0.2403	25	0.83	012.00	0
012	Starch (%)	2196	10.06	0.0000	8.876	1.311	0.2403	25	0.90	012.00	0
012	Starch (%)	2089	10.46	0.1500	8.876	1.311	0.2403	25	1.20	012.01	0
012	Starch (%)	0885	12.19	0.3373	8.876	1.311	0.2403	25	2.52	012.01	0
012	Starch (%)	0852	13.26	0.0300	8.876	1.311	0.2403	25	3.34	012.11	0
012	Starch (%)	0720	5.105	2.070	8.876	1.311	0.2403	25	-2.88	012.11	1
013	Fat, Acid Pretreat (%)	0720	3.440	0.1600	4.775	0.7473	0.1340	47	-1.79	013.10	0
013	Fat, Acid Pretreat (%)	2073	3.515	0.0100	4.775	0.7473	0.1340	47	-1.69	013.00	0
013	Fat, Acid Pretreat (%)	0651	3.585	0.0000	4.775	0.7473	0.1340	47	-1.59	013.13	0
013	Fat, Acid Pretreat (%)	0036	3.725	0.0218	4.775	0.7473	0.1340	47	-1.41	013.10	0
013	Fat, Acid Pretreat (%)	0689	3.750	0.1000	4.775	0.7473	0.1340	47	-1.37	013.00	0
013	Fat, Acid Pretreat (%)	2345	3.850	0.3000	4.775	0.7473	0.1340	47	-1.24	013.00	0
013	Fat, Acid Pretreat (%)	2268	4.070	0.1000	4.775	0.7473	0.1340	47	-0.94	013.10	0
013	Fat, Acid Pretreat (%)	0083	4.100	0.0600	4.775	0.7473	0.1340	47	-0.90	013.00	0
013	Fat, Acid Pretreat (%)	0843	4.115	0.1700	4.775	0.7473	0.1340	47	-0.88	013.13	0
013	Fat, Acid Pretreat (%)	0643	4.134	0.0560	4.775	0.7473	0.1340	47	-0.86	013.02	0
013	Fat, Acid Pretreat (%)	0242	4.170	0.1200	4.775	0.7473	0.1340	47	-0.81	013.00	0
013	Fat, Acid Pretreat (%)	0354	4.225	0.0300	4.775	0.7473	0.1340	47	-0.74	013.02	0
013	Fat, Acid Pretreat (%)	0610	4.275	0.0500	4.775	0.7473	0.1340	47	-0.67	013.10	0
013	Fat, Acid Pretreat (%)	2319	4.300	0.0000	4.775	0.7473	0.1340	47	-0.64	013.00	0
013	Fat, Acid Pretreat (%)	2196	4.340	0.0000	4.775	0.7473	0.1340	47	-0.58	013.10	0
013	Fat, Acid Pretreat (%)	0910	4.400	0.2400	4.775	0.7473	0.1340	47	-0.50	013.00	0
013	Fat, Acid Pretreat (%)	0948	4.475	0.0900	4.775	0.7473	0.1340	47	-0.40	013.00	0
013	Fat, Acid Pretreat (%)	0353	4.580	0.4400	4.775	0.7473	0.1340	47	-0.26	013.10	0
013	Fat, Acid Pretreat (%)	0682	4.600	0.0000	4.775	0.7473	0.1340	47	-0.23	013.02	0
013	Fat, Acid Pretreat (%)	0881	4.634	0.0010	4.775	0.7473	0.1340	47	-0.19	013.00	0
013	Fat, Acid Pretreat (%)	0676	4.725	0.1900	4.775	0.7473	0.1340	47	-0.07	013.02	0
013	Fat, Acid Pretreat (%)	0914	4.740	0.1900	4.775	0.7473	0.1340	47	-0.05	013.02	0
013	Fat, Acid Pretreat (%)	0884	4.750	0.1000	4.775	0.7473	0.1340	47	-0.03	013.00	0
013	Fat, Acid Pretreat (%)	0921	4.796	0.0130	4.775	0.7473	0.1340	47	0.03	013.02	0
013	Fat, Acid Pretreat (%)	0098	4.805	0.0300	4.775	0.7473	0.1340	47	0.04	013.02	0
013	Fat, Acid Pretreat (%)	0853	4.850	0.1000	4.775	0.7473	0.1340	47	0.10	013.02	0
013	Fat, Acid Pretreat (%)	0541	4.945	0.1500	4.775	0.7473	0.1340	47	0.23	013.00	0
013	Fat, Acid Pretreat (%)	0652	4.950	0.3000	4.775	0.7473	0.1340	47	0.23	013.00	0
013	Fat, Acid Pretreat (%)	0939	5.030	0.3200	4.775	0.7473	0.1340	47	0.34	013.13	0
013	Fat, Acid Pretreat (%)	0208	5.040	0.0400	4.775	0.7473	0.1340	47	0.35	013.02	0
013	Fat, Acid Pretreat (%)	0870	5.062	0.0038	4.775	0.7473	0.1340	47	0.38	013.02	0
013	Fat, Acid Pretreat (%)	0297	5.105	0.4500	4.775	0.7473	0.1340	47	0.44	013.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
013	Fat, Acid Pretreat (%)	0968	5.160	0.5400	4.775	0.7473	0.1340	47	0.51	013.13	0
013	Fat, Acid Pretreat (%)	0148	5.170	0.1800	4.775	0.7473	0.1340	47	0.53	013.02	0
013	Fat, Acid Pretreat (%)	2328	5.370	0.0400	4.775	0.7473	0.1340	47	0.80	013.02	0
013	Fat, Acid Pretreat (%)	0581	5.380	0.1000	4.775	0.7473	0.1340	47	0.81	013.13	0
013	Fat, Acid Pretreat (%)	2076	5.427	0.1653	4.775	0.7473	0.1340	47	0.87	013.00	0
013	Fat, Acid Pretreat (%)	0202	5.430	0.0800	4.775	0.7473	0.1340	47	0.88	013.00	0
013	Fat, Acid Pretreat (%)	0407	5.450	0.0200	4.775	0.7473	0.1340	47	0.90	013.02	0
013	Fat, Acid Pretreat (%)	0553	5.455	0.0500	4.775	0.7473	0.1340	47	0.91	013.02	0
013	Fat, Acid Pretreat (%)	2150	5.470	0.2600	4.775	0.7473	0.1340	47	0.93	013.00	0
013	Fat, Acid Pretreat (%)	0675	5.485	0.3300	4.775	0.7473	0.1340	47	0.95	013.02	0
013	Fat, Acid Pretreat (%)	0171	5.700	0.1200	4.775	0.7473	0.1340	47	1.24	013.02	0
013	Fat, Acid Pretreat (%)	0309	5.814	0.0530	4.775	0.7473	0.1340	47	1.39	013.00	0
013	Fat, Acid Pretreat (%)	0407	5.845	0.2051	4.775	0.7473	0.1340	47	1.43	013.13	0
013	Fat, Acid Pretreat (%)	0265	5.885	0.0500	4.775	0.7473	0.1340	47	1.48	013.13	0
013	Fat, Acid Pretreat (%)	2259	5.996	0.2680	4.775	0.7473	0.1340	47	1.63	013.02	0
015	Aluminum (ppm)	0037	197.4	21.10	643.6	150.5	10.42	16	-2.97	015.42	0
015	Aluminum (ppm)	0297	255.5	5.000	643.6	150.5	10.42	16	-2.58	015.43	0
015	Aluminum (ppm)	0049	455.3	9.580	643.6	150.5	10.42	16	-1.25	015.41	0
015	Aluminum (ppm)	0123	587.5	17.00	643.6	150.5	10.42	16	-0.37	015.41	0
015	Aluminum (ppm)	0171	590.0	0.0000	643.6	150.5	10.42	16	-0.36	015.41	0
015	Aluminum (ppm)	0425	596.7	10.70	643.6	150.5	10.42	16	-0.31	015.43	0
015	Aluminum (ppm)	0407	614.4	3.048	643.6	150.5	10.42	16	-0.19	015.43	0
015	Aluminum (ppm)	0407	636.2	9.076	643.6	150.5	10.42	16	-0.05	015.41	0
015	Aluminum (ppm)	0345	640.5	12.60	643.6	150.5	10.42	16	-0.02	015.43	0
015	Aluminum (ppm)	0520	682.5	15.00	643.6	150.5	10.42	16	0.26	015.41	0
015	Aluminum (ppm)	0510	697.5	3.000	643.6	150.5	10.42	16	0.36	015.43	0
015	Aluminum (ppm)	0407	752.8	25.55	643.6	150.5	10.42	16	0.73	015.53	0
015	Aluminum (ppm)	0553	787.5	1.000	643.6	150.5	10.42	16	0.96	015.53	0
015	Aluminum (ppm)	0169	789.5	5.000	643.6	150.5	10.42	16	0.97	015.43	0
015	Aluminum (ppm)	0353	812.5	21.00	643.6	150.5	10.42	16	1.12	015.43	0
015	Aluminum (ppm)	0939	819.0	8.000	643.6	150.5	10.42	16	1.17	015.53	0
017	Boron (ppm)	0297	5.000	0.0000	20.84	2.419	0.7556	16	-6.55	017.43	0
017	Boron (ppm)	0123	13.15	0.1000	20.84	2.419	0.7556	16	-3.18	017.44	0
017	Boron (ppm)	0358	18.79	0.6700	20.84	2.419	0.7556	16	-0.85	017.41	0
017	Boron (ppm)	0345	18.99	0.1800	20.84	2.419	0.7556	16	-0.77	017.43	0
017	Boron (ppm)	0226	19.83	0.2400	20.84	2.419	0.7556	16	-0.42	017.41	0
017	Boron (ppm)	0407	20.22	0.4898	20.84	2.419	0.7556	16	-0.26	017.41	0
017	Boron (ppm)	0294	20.30	0.4000	20.84	2.419	0.7556	16	-0.22	017.42	0
017	Boron (ppm)	0049	20.91	0.8900	20.84	2.419	0.7556	16	0.03	017.41	0
017	Boron (ppm)	0037	20.95	0.3000	20.84	2.419	0.7556	16	0.04	017.42	0
017	Boron (ppm)	0045	21.30	0.0000	20.84	2.419	0.7556	16	0.19	017.42	0
017	Boron (ppm)	0407	21.73	2.522	20.84	2.419	0.7556	16	0.37	017.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
017	Boron (ppm)	0265	22.10	1.800	20.84	2.419	0.7556	16	0.52	017.42	0
017	Boron (ppm)	0353	22.50	1.000	20.84	2.419	0.7556	16	0.68	017.43	0
017	Boron (ppm)	0407	23.49	2.198	20.84	2.419	0.7556	16	1.09	017.53	0
017	Boron (ppm)	0510	23.50	1.000	20.84	2.419	0.7556	16	1.10	017.43	0
017	Boron (ppm)	0083	25.45	0.3000	20.84	2.419	0.7556	16	1.90	017.43	0
017	Boron (ppm)	0693	16.42	5.419	20.84	2.419	0.7556	16	-1.83	017.42	1
019	Calcium (%)	2129	0.9185	0.0290	1.129	0.0700	0.0291	113	-3.01	019.42	0
019	Calcium (%)	0242	0.9750	0.0100	1.129	0.0700	0.0291	113	-2.20	019.99	0
019	Calcium (%)	2006	1.007	0.0540	1.129	0.0700	0.0291	113	-1.74	019.00	0
019	Calcium (%)	2109	1.015	0.0670	1.129	0.0700	0.0291	113	-1.64	019.00	0
019	Calcium (%)	0006	1.020	0.0520	1.129	0.0700	0.0291	113	-1.56	019.43	0
019	Calcium (%)	0036	1.025	0.0038	1.129	0.0700	0.0291	113	-1.49	019.43	0
019	Calcium (%)	0910	1.030	0.0600	1.129	0.0700	0.0291	113	-1.42	019.41	0
019	Calcium (%)	0914	1.030	0.0200	1.129	0.0700	0.0291	113	-1.42	019.43	0
019	Calcium (%)	0171	1.035	0.0100	1.129	0.0700	0.0291	113	-1.34	019.41	0
019	Calcium (%)	2006	1.036	0.0210	1.129	0.0700	0.0291	113	-1.34	019.09	0
019	Calcium (%)	2196	1.040	0.0000	1.129	0.0700	0.0291	113	-1.27	019.08	0
019	Calcium (%)	0142	1.047	0.0242	1.129	0.0700	0.0291	113	-1.18	019.31	0
019	Calcium (%)	2009	1.049	0.0056	1.129	0.0700	0.0291	113	-1.15	019.08	0
019	Calcium (%)	2196	1.050	0.0000	1.129	0.0700	0.0291	113	-1.13	019.31	0
019	Calcium (%)	0674	1.050	0.0000	1.129	0.0700	0.0291	113	-1.13	019.41	0
019	Calcium (%)	0870	1.054	0.0220	1.129	0.0700	0.0291	113	-1.07	019.42	0
019	Calcium (%)	0169	1.055	0.0100	1.129	0.0700	0.0291	113	-1.06	019.32	0
019	Calcium (%)	0968	1.059	0.0010	1.129	0.0700	0.0291	113	-1.01	019.43	0
019	Calcium (%)	0529	1.060	0.0800	1.129	0.0700	0.0291	113	-0.99	019.31	0
019	Calcium (%)	0511	1.065	0.0300	1.129	0.0700	0.0291	113	-0.91	019.43	0
019	Calcium (%)	2193	1.070	0.0400	1.129	0.0700	0.0291	113	-0.84	019.08	0
019	Calcium (%)	0004	1.070	0.0200	1.129	0.0700	0.0291	113	-0.84	019.41	0
019	Calcium (%)	0510	1.070	0.0000	1.129	0.0700	0.0291	113	-0.84	019.43	0
019	Calcium (%)	2340	1.074	0.0020	1.129	0.0700	0.0291	113	-0.79	019.08	0
019	Calcium (%)	0066	1.075	0.0130	1.129	0.0700	0.0291	113	-0.78	019.41	0
019	Calcium (%)	0263	1.077	0.0010	1.129	0.0700	0.0291	113	-0.75	019.42	0
019	Calcium (%)	2058	1.079	0.0630	1.129	0.0700	0.0291	113	-0.72	019.43	0
019	Calcium (%)	2319	1.080	0.0400	1.129	0.0700	0.0291	113	-0.70	019.42	0
019	Calcium (%)	0939	1.080	0.0200	1.129	0.0700	0.0291	113	-0.70	019.53	0
019	Calcium (%)	0019	1.085	0.0100	1.129	0.0700	0.0291	113	-0.63	019.41	0
019	Calcium (%)	0692	1.085	0.0100	1.129	0.0700	0.0291	113	-0.63	019.42	0
019	Calcium (%)	0536	1.090	0.0000	1.129	0.0700	0.0291	113	-0.56	019.31	0
019	Calcium (%)	2192	1.090	0.0000	1.129	0.0700	0.0291	113	-0.56	019.43	0
019	Calcium (%)	0553	1.090	0.0400	1.129	0.0700	0.0291	113	-0.56	019.53	0
019	Calcium (%)	0345	1.092	0.0160	1.129	0.0700	0.0291	113	-0.53	019.43	0
019	Calcium (%)	0881	1.094	0.0550	1.129	0.0700	0.0291	113	-0.51	019.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
019	Calcium (%)	0047	1.095	0.0700	1.129	0.0700	0.0291	113	-0.49	019.52	0
019	Calcium (%)	0208	1.097	0.0910	1.129	0.0700	0.0291	113	-0.46	019.41	0
019	Calcium (%)	0598	1.097	0.0005	1.129	0.0700	0.0291	113	-0.45	019.41	0
019	Calcium (%)	0358	1.100	0.0200	1.129	0.0700	0.0291	113	-0.41	019.41	0
019	Calcium (%)	0038	1.104	0.0690	1.129	0.0700	0.0291	113	-0.36	019.43	0
019	Calcium (%)	0049	1.105	0.0100	1.129	0.0700	0.0291	113	-0.34	019.41	0
019	Calcium (%)	0278	1.105	0.0100	1.129	0.0700	0.0291	113	-0.34	019.42	0
019	Calcium (%)	0018	1.110	0.0400	1.129	0.0700	0.0291	113	-0.27	019.41	0
019	Calcium (%)	0619	1.110	0.0400	1.129	0.0700	0.0291	113	-0.27	019.41	0
019	Calcium (%)	0848	1.110	0.0000	1.129	0.0700	0.0291	113	-0.27	019.41	0
019	Calcium (%)	0017	1.110	0.0009	1.129	0.0700	0.0291	113	-0.27	019.43	0
019	Calcium (%)	0265	1.115	0.0100	1.129	0.0700	0.0291	113	-0.20	019.42	0
019	Calcium (%)	0689	1.120	0.0000	1.129	0.0700	0.0291	113	-0.13	019.08	0
019	Calcium (%)	0226	1.120	0.0000	1.129	0.0700	0.0291	113	-0.13	019.41	0
019	Calcium (%)	0354	1.120	0.0200	1.129	0.0700	0.0291	113	-0.13	019.41	0
019	Calcium (%)	0033	1.120	0.0200	1.129	0.0700	0.0291	113	-0.13	019.43	0
019	Calcium (%)	0407	1.122	0.0314	1.129	0.0700	0.0291	113	-0.10	019.43	0
019	Calcium (%)	0154	1.125	0.0872	1.129	0.0700	0.0291	113	-0.06	019.52	0
019	Calcium (%)	0098	1.130	0.0400	1.129	0.0700	0.0291	113	0.01	019.41	0
019	Calcium (%)	0366	1.130	0.0600	1.129	0.0700	0.0291	113	0.01	019.42	0
019	Calcium (%)	0505	1.130	0.0600	1.129	0.0700	0.0291	113	0.01	019.43	0
019	Calcium (%)	0208	1.131	0.0890	1.129	0.0700	0.0291	113	0.02	019.31	0
019	Calcium (%)	2190	1.131	0.0006	1.129	0.0700	0.0291	113	0.03	019.08	0
019	Calcium (%)	0407	1.134	0.0416	1.129	0.0700	0.0291	113	0.07	019.41	0
019	Calcium (%)	2089	1.135	0.0100	1.129	0.0700	0.0291	113	0.09	019.43	0
019	Calcium (%)	2259	1.137	0.0170	1.129	0.0700	0.0291	113	0.11	019.41	0
019	Calcium (%)	2325	1.140	0.0200	1.129	0.0700	0.0291	113	0.16	019.31	0
019	Calcium (%)	0693	1.143	0.0250	1.129	0.0700	0.0291	113	0.19	019.42	0
019	Calcium (%)	0885	1.145	0.0048	1.129	0.0700	0.0291	113	0.23	019.08	0
019	Calcium (%)	0011	1.145	0.0700	1.129	0.0700	0.0291	113	0.23	019.41	0
019	Calcium (%)	0123	1.145	0.0100	1.129	0.0700	0.0291	113	0.23	019.41	0
019	Calcium (%)	0035	1.145	0.0100	1.129	0.0700	0.0291	113	0.23	019.42	0
019	Calcium (%)	0010	1.145	0.0100	1.129	0.0700	0.0291	113	0.23	019.43	0
019	Calcium (%)	0009	1.146	0.0588	1.129	0.0700	0.0291	113	0.25	019.42	0
019	Calcium (%)	0870	1.150	0.0490	1.129	0.0700	0.0291	113	0.29	019.43	0
019	Calcium (%)	2337	1.150	0.0000	1.129	0.0700	0.0291	113	0.30	019.08	0
019	Calcium (%)	0687	1.150	0.1000	1.129	0.0700	0.0291	113	0.30	019.31	0
019	Calcium (%)	0045	1.150	0.0600	1.129	0.0700	0.0291	113	0.30	019.42	0
019	Calcium (%)	0651	1.154	0.0259	1.129	0.0700	0.0291	113	0.36	019.31	0
019	Calcium (%)	0144	1.155	0.0500	1.129	0.0700	0.0291	113	0.37	019.41	0
019	Calcium (%)	0083	1.155	0.0100	1.129	0.0700	0.0291	113	0.37	019.43	0
019	Calcium (%)	0612	1.160	0.0000	1.129	0.0700	0.0291	113	0.44	019.33	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
019	Calcium (%)	0610	1.160	0.0000	1.129	0.0700	0.0291	113	0.44	019.43	0
019	Calcium (%)	0682	1.160	0.0000	1.129	0.0700	0.0291	113	0.44	019.43	0
019	Calcium (%)	0202	1.165	0.0300	1.129	0.0700	0.0291	113	0.51	019.43	0
019	Calcium (%)	0027	1.167	0.0175	1.129	0.0700	0.0291	113	0.54	019.43	0
019	Calcium (%)	0148	1.170	0.0000	1.129	0.0700	0.0291	113	0.59	019.41	0
019	Calcium (%)	0425	1.170	0.0800	1.129	0.0700	0.0291	113	0.59	019.43	0
019	Calcium (%)	0675	1.170	0.0200	1.129	0.0700	0.0291	113	0.59	019.43	0
019	Calcium (%)	0572	1.170	0.0200	1.129	0.0700	0.0291	113	0.59	019.53	0
019	Calcium (%)	0100	1.170	0.0000	1.129	0.0700	0.0291	113	0.59	019.99	0
019	Calcium (%)	0098	1.175	0.0500	1.129	0.0700	0.0291	113	0.66	019.43	0
019	Calcium (%)	0108	1.175	0.1500	1.129	0.0700	0.0291	113	0.66	019.31	0
019	Calcium (%)	0001	1.185	0.0150	1.129	0.0700	0.0291	113	0.79	019.31	0
019	Calcium (%)	0357	1.186	0.0203	1.129	0.0700	0.0291	113	0.81	019.42	0
019	Calcium (%)	0038	1.188	0.0110	1.129	0.0700	0.0291	113	0.84	019.31	0
019	Calcium (%)	0037	1.188	0.0560	1.129	0.0700	0.0291	113	0.84	019.42	0
019	Calcium (%)	0035	1.190	0.0200	1.129	0.0700	0.0291	113	0.87	019.43	0
019	Calcium (%)	0353	1.190	0.0200	1.129	0.0700	0.0291	113	0.87	019.43	0
019	Calcium (%)	0042	1.195	0.0100	1.129	0.0700	0.0291	113	0.94	019.42	0
019	Calcium (%)	0646	1.200	0.1000	1.129	0.0700	0.0291	113	1.01	019.31	0
019	Calcium (%)	2342	1.200	0.0600	1.129	0.0700	0.0291	113	1.01	019.44	0
019	Calcium (%)	0720	1.225	0.0300	1.129	0.0700	0.0291	113	1.37	019.41	0
019	Calcium (%)	0294	1.225	0.0100	1.129	0.0700	0.0291	113	1.37	019.42	0
019	Calcium (%)	2302	1.230	0.0000	1.129	0.0700	0.0291	113	1.44	019.00	0
019	Calcium (%)	2302	1.230	0.0000	1.129	0.0700	0.0291	113	1.44	019.99	0
019	Calcium (%)	0563	1.239	0.0237	1.129	0.0700	0.0291	113	1.57	019.31	0
019	Calcium (%)	0036	1.248	0.0348	1.129	0.0700	0.0291	113	1.70	019.03	0
019	Calcium (%)	2326	1.250	0.1000	1.129	0.0700	0.0291	113	1.73	019.43	0
019	Calcium (%)	2188	1.270	0.0000	1.129	0.0700	0.0291	113	2.02	019.08	0
019	Calcium (%)	2322	1.273	0.0150	1.129	0.0700	0.0291	113	2.05	019.43	0
019	Calcium (%)	0190	1.275	0.0100	1.129	0.0700	0.0291	113	2.09	019.42	0
019	Calcium (%)	0620	1.278	0.0136	1.129	0.0700	0.0291	113	2.13	019.31	0
019	Calcium (%)	0520	1.280	0.0400	1.129	0.0700	0.0291	113	2.16	019.41	0
019	Calcium (%)	0726	1.305	0.0500	1.129	0.0700	0.0291	113	2.52	019.42	0
019	Calcium (%)	0884	1.375	0.0500	1.129	0.0700	0.0291	113	3.52	019.31	0
019	Calcium (%)	0407	1.424	0.0637	1.129	0.0700	0.0291	113	4.21	019.53	0
019	Calcium (%)	0852	0.4000	0.2000	1.129	0.0700	0.0291	113	-10.42	019.08	1
019	Calcium (%)	0297	4.800	0.3200	1.129	0.0700	0.0291	113	52.48	019.43	2
021	Cobalt (ppm)	0047	3.075	0.5900	8.065	2.698	0.5212	23	-1.85	021.52	0
021	Cobalt (ppm)	0910	3.990	0.6000	8.065	2.698	0.5212	23	-1.51	021.52	0
021	Cobalt (ppm)	0619	4.850	0.1000	8.065	2.698	0.5212	23	-1.19	021.41	0
021	Cobalt (ppm)	0171	5.290	0.2200	8.065	2.698	0.5212	23	-1.03	021.41	0
021	Cobalt (ppm)	0160	5.750	0.1000	8.065	2.698	0.5212	23	-0.86	021.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
021	Cobalt (ppm)	0123	6.275	0.4900	8.065	2.698	0.5212	23	-0.66	021.41	0
021	Cobalt (ppm)	0345	6.378	0.0650	8.065	2.698	0.5212	23	-0.63	021.43	0
021	Cobalt (ppm)	0045	6.455	1.630	8.065	2.698	0.5212	23	-0.60	021.42	0
021	Cobalt (ppm)	0693	6.922	1.079	8.065	2.698	0.5212	23	-0.42	021.42	0
021	Cobalt (ppm)	0066	7.403	0.0780	8.065	2.698	0.5212	23	-0.25	021.41	0
021	Cobalt (ppm)	0407	8.052	1.145	8.065	2.698	0.5212	23	0.00	021.43	0
021	Cobalt (ppm)	0407	8.352	0.0008	8.065	2.698	0.5212	23	0.11	021.41	0
021	Cobalt (ppm)	0033	8.780	0.4600	8.065	2.698	0.5212	23	0.26	021.53	0
021	Cobalt (ppm)	0038	8.966	0.7350	8.065	2.698	0.5212	23	0.33	021.43	0
021	Cobalt (ppm)	0939	9.220	0.2200	8.065	2.698	0.5212	23	0.43	021.53	0
021	Cobalt (ppm)	0553	9.595	0.3300	8.065	2.698	0.5212	23	0.57	021.53	0
021	Cobalt (ppm)	0148	9.830	0.3600	8.065	2.698	0.5212	23	0.65	021.41	0
021	Cobalt (ppm)	2326	10.14	0.0700	8.065	2.698	0.5212	23	0.77	021.43	0
021	Cobalt (ppm)	0407	10.44	0.7141	8.065	2.698	0.5212	23	0.88	021.53	0
021	Cobalt (ppm)	0353	10.86	2.601	8.065	2.698	0.5212	23	1.04	021.43	0
021	Cobalt (ppm)	0169	10.95	0.3000	8.065	2.698	0.5212	23	1.07	021.43	0
021	Cobalt (ppm)	0682	11.00	0.0000	8.065	2.698	0.5212	23	1.09	021.43	0
021	Cobalt (ppm)	0610	11.95	0.1000	8.065	2.698	0.5212	23	1.44	021.53	0
021	Cobalt (ppm)	0510	933.0	12.00	8.065	2.698	0.5212	23	342.79	021.43	2
022	Copper (ppm)	0297	11.00	0.0000	44.64	9.547	1.844	81	-3.52	022.43	0
022	Copper (ppm)	0049	16.30	0.4900	44.64	9.547	1.844	81	-2.97	022.41	0
022	Copper (ppm)	0598	18.68	0.3307	44.64	9.547	1.844	81	-2.72	022.41	0
022	Copper (ppm)	0004	23.47	1.760	44.64	9.547	1.844	81	-2.22	022.41	0
022	Copper (ppm)	0536	24.81	0.0700	44.64	9.547	1.844	81	-2.08	022.31	0
022	Copper (ppm)	0848	26.55	0.2300	44.64	9.547	1.844	81	-1.90	022.41	0
022	Copper (ppm)	0619	28.80	0.2000	44.64	9.547	1.844	81	-1.66	022.41	0
022	Copper (ppm)	0019	29.25	1.060	44.64	9.547	1.844	81	-1.61	022.41	0
022	Copper (ppm)	0910	31.50	1.000	44.64	9.547	1.844	81	-1.38	022.41	0
022	Copper (ppm)	0358	31.74	0.6500	44.64	9.547	1.844	81	-1.35	022.41	0
022	Copper (ppm)	0171	32.19	0.2200	44.64	9.547	1.844	81	-1.30	022.41	0
022	Copper (ppm)	0720	33.13	0.1810	44.64	9.547	1.844	81	-1.21	022.41	0
022	Copper (ppm)	0066	33.21	0.9500	44.64	9.547	1.844	81	-1.20	022.44	0
022	Copper (ppm)	0407	33.57	0.7522	44.64	9.547	1.844	81	-1.16	022.41	0
022	Copper (ppm)	0098	33.93	2.490	44.64	9.547	1.844	81	-1.12	022.41	0
022	Copper (ppm)	2196	34.95	0.0000	44.64	9.547	1.844	81	-1.02	022.31	0
022	Copper (ppm)	0563	35.23	1.700	44.64	9.547	1.844	81	-0.99	022.31	0
022	Copper (ppm)	0123	35.85	0.1000	44.64	9.547	1.844	81	-0.92	022.41	0
022	Copper (ppm)	0148	36.77	5.390	44.64	9.547	1.844	81	-0.83	022.41	0
022	Copper (ppm)	0520	37.00	6.000	44.64	9.547	1.844	81	-0.80	022.41	0
022	Copper (ppm)	0208	37.25	4.900	44.64	9.547	1.844	81	-0.77	022.31	0
022	Copper (ppm)	0674	37.28	0.0000	44.64	9.547	1.844	81	-0.77	022.41	0
022	Copper (ppm)	0001	37.60	0.5700	44.64	9.547	1.844	81	-0.74	022.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
022	Copper (ppm)	0689	38.65	1.500	44.64	9.547	1.844	81	-0.63	022.31	0
022	Copper (ppm)	0208	38.77	4.880	44.64	9.547	1.844	81	-0.62	022.41	0
022	Copper (ppm)	2188	38.77	0.0000	44.64	9.547	1.844	81	-0.62	022.31	0
022	Copper (ppm)	0529	39.05	1.700	44.64	9.547	1.844	81	-0.59	022.31	0
022	Copper (ppm)	0226	40.45	0.6700	44.64	9.547	1.844	81	-0.44	022.41	0
022	Copper (ppm)	0047	40.64	2.490	44.64	9.547	1.844	81	-0.42	022.52	0
022	Copper (ppm)	0242	44.50	1.000	44.64	9.547	1.844	81	-0.01	022.99	0
022	Copper (ppm)	2322	44.53	0.3600	44.64	9.547	1.844	81	-0.01	022.43	0
022	Copper (ppm)	0263	44.61	0.6970	44.64	9.547	1.844	81	0.00	022.42	0
022	Copper (ppm)	0407	45.28	0.1806	44.64	9.547	1.844	81	0.07	022.43	0
022	Copper (ppm)	0033	45.45	1.700	44.64	9.547	1.844	81	0.08	022.53	0
022	Copper (ppm)	0511	45.50	1.000	44.64	9.547	1.844	81	0.09	022.43	0
022	Copper (ppm)	0726	45.79	0.1600	44.64	9.547	1.844	81	0.12	022.42	0
022	Copper (ppm)	0553	46.20	0.8000	44.64	9.547	1.844	81	0.16	022.53	0
022	Copper (ppm)	0968	46.81	0.1970	44.64	9.547	1.844	81	0.23	022.43	0
022	Copper (ppm)	0939	47.86	2.740	44.64	9.547	1.844	81	0.34	022.53	0
022	Copper (ppm)	0100	48.00	0.0000	44.64	9.547	1.844	81	0.35	022.99	0
022	Copper (ppm)	2319	48.15	0.1000	44.64	9.547	1.844	81	0.37	022.42	0
022	Copper (ppm)	0017	48.16	3.744	44.64	9.547	1.844	81	0.37	022.43	0
022	Copper (ppm)	0202	48.25	0.7000	44.64	9.547	1.844	81	0.38	022.43	0
022	Copper (ppm)	0278	48.30	2.200	44.64	9.547	1.844	81	0.38	022.42	0
022	Copper (ppm)	0610	48.40	0.6000	44.64	9.547	1.844	81	0.39	022.43	0
022	Copper (ppm)	0510	48.50	1.000	44.64	9.547	1.844	81	0.40	022.43	0
022	Copper (ppm)	0045	48.70	0.2000	44.64	9.547	1.844	81	0.42	022.42	0
022	Copper (ppm)	0345	48.75	0.7000	44.64	9.547	1.844	81	0.43	022.43	0
022	Copper (ppm)	0035	49.00	0.0000	44.64	9.547	1.844	81	0.46	022.43	0
022	Copper (ppm)	0033	49.10	1.400	44.64	9.547	1.844	81	0.47	022.43	0
022	Copper (ppm)	0265	49.15	3.900	44.64	9.547	1.844	81	0.47	022.42	0
022	Copper (ppm)	0357	49.32	0.1700	44.64	9.547	1.844	81	0.49	022.42	0
022	Copper (ppm)	2058	49.50	2.870	44.64	9.547	1.844	81	0.51	022.43	0
022	Copper (ppm)	0010	49.50	3.000	44.64	9.547	1.844	81	0.51	022.33	0
022	Copper (ppm)	0035	49.50	1.000	44.64	9.547	1.844	81	0.51	022.42	0
022	Copper (ppm)	0354	49.57	4.310	44.64	9.547	1.844	81	0.52	022.41	0
022	Copper (ppm)	0294	49.75	1.700	44.64	9.547	1.844	81	0.53	022.42	0
022	Copper (ppm)	0651	49.89	3.807	44.64	9.547	1.844	81	0.55	022.31	0
022	Copper (ppm)	0098	49.90	1.180	44.64	9.547	1.844	81	0.55	022.43	0
022	Copper (ppm)	2089	49.99	3.120	44.64	9.547	1.844	81	0.56	022.43	0
022	Copper (ppm)	0083	50.00	0.0000	44.64	9.547	1.844	81	0.56	022.43	0
022	Copper (ppm)	0870	50.88	1.070	44.64	9.547	1.844	81	0.65	022.42	0
022	Copper (ppm)	0366	51.00	2.000	44.64	9.547	1.844	81	0.67	022.42	0
022	Copper (ppm)	2342	51.00	2.400	44.64	9.547	1.844	81	0.67	022.44	0
022	Copper (ppm)	2129	51.25	3.300	44.64	9.547	1.844	81	0.69	022.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
022	Copper (ppm)	0190	52.00	0.4100	44.64	9.547	1.844	81	0.77	022.42	0
022	Copper (ppm)	0682	52.10	0.0000	44.64	9.547	1.844	81	0.78	022.43	0
022	Copper (ppm)	0009	52.50	3.000	44.64	9.547	1.844	81	0.82	022.42	0
022	Copper (ppm)	0692	52.75	4.900	44.64	9.547	1.844	81	0.85	022.42	0
022	Copper (ppm)	2326	53.33	2.930	44.64	9.547	1.844	81	0.91	022.43	0
022	Copper (ppm)	0169	53.55	0.1000	44.64	9.547	1.844	81	0.93	022.43	0
022	Copper (ppm)	2192	53.75	3.300	44.64	9.547	1.844	81	0.95	022.43	0
022	Copper (ppm)	0425	53.96	5.230	44.64	9.547	1.844	81	0.98	022.43	0
022	Copper (ppm)	0572	54.00	10.20	44.64	9.547	1.844	81	0.98	022.53	0
022	Copper (ppm)	0675	55.27	0.5600	44.64	9.547	1.844	81	1.11	022.43	0
022	Copper (ppm)	0353	56.00	2.000	44.64	9.547	1.844	81	1.19	022.43	0
022	Copper (ppm)	0693	56.14	8.469	44.64	9.547	1.844	81	1.20	022.42	0
022	Copper (ppm)	0870	56.56	3.030	44.64	9.547	1.844	81	1.25	022.43	0
022	Copper (ppm)	0407	57.66	0.4087	44.64	9.547	1.844	81	1.36	022.53	0
022	Copper (ppm)	0027	57.87	10.16	44.64	9.547	1.844	81	1.39	022.43	0
022	Copper (ppm)	0042	58.55	1.100	44.64	9.547	1.844	81	1.46	022.42	0
022	Copper (ppm)	0038	20.07	13.56	44.64	9.547	1.844	81	-2.57	022.32	1
022	Copper (ppm)	0037	57.20	17.60	44.64	9.547	1.844	81	1.32	022.42	1
024	Iodine (ppm)	0160	4.385	1.370				1		024.99	0
025	Iron (ppm)	2322	118.5	2.390	700.3	131.6	22.31	61	-4.42	025.43	0
025	Iron (ppm)	0726	253.2	17.60	700.3	131.6	22.31	61	-3.40	025.42	0
025	Iron (ppm)	0037	256.5	12.30	700.3	131.6	22.31	61	-3.37	025.42	0
025	Iron (ppm)	0265	367.0	12.00	700.3	131.6	22.31	61	-2.53	025.42	0
025	Iron (ppm)	2342	378.0	16.00	700.3	131.6	22.31	61	-2.45	025.99	0
025	Iron (ppm)	0294	444.6	5.040	700.3	131.6	22.31	61	-1.94	025.42	0
025	Iron (ppm)	0425	471.1	101.2	700.3	131.6	22.31	61	-1.74	025.43	0
025	Iron (ppm)	0297	494.0	10.00	700.3	131.6	22.31	61	-1.57	025.43	0
025	Iron (ppm)	2319	515.0	34.00	700.3	131.6	22.31	61	-1.41	025.42	0
025	Iron (ppm)	2089	543.9	58.12	700.3	131.6	22.31	61	-1.19	025.43	0
025	Iron (ppm)	0689	545.5	1.400	700.3	131.6	22.31	61	-1.18	025.31	0
025	Iron (ppm)	0692	557.5	5.000	700.3	131.6	22.31	61	-1.08	025.42	0
025	Iron (ppm)	0720	570.8	3.700	700.3	131.6	22.31	61	-0.98	025.41	0
025	Iron (ppm)	0354	579.6	17.49	700.3	131.6	22.31	61	-0.92	025.41	0
025	Iron (ppm)	0278	610.5	41.00	700.3	131.6	22.31	61	-0.68	025.42	0
025	Iron (ppm)	0598	631.4	22.01	700.3	131.6	22.31	61	-0.52	025.41	0
025	Iron (ppm)	0536	652.9	12.47	700.3	131.6	22.31	61	-0.36	025.31	0
025	Iron (ppm)	0083	661.5	1.000	700.3	131.6	22.31	61	-0.29	025.43	0
025	Iron (ppm)	0242	662.0	18.00	700.3	131.6	22.31	61	-0.29	025.99	0
025	Iron (ppm)	0045	666.5	7.000	700.3	131.6	22.31	61	-0.26	025.42	0
025	Iron (ppm)	0226	676.7	13.52	700.3	131.6	22.31	61	-0.18	025.41	0
025	Iron (ppm)	0870	682.6	10.20	700.3	131.6	22.31	61	-0.13	025.42	0
025	Iron (ppm)	0366	688.5	3.000	700.3	131.6	22.31	61	-0.09	025.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
025	Iron (ppm)	0693	691.9	53.55	700.3	131.6	22.31	61	-0.06	025.42	0
025	Iron (ppm)	0968	697.2	2.011	700.3	131.6	22.31	61	-0.02	025.43	0
025	Iron (ppm)	0345	699.5	19.00	700.3	131.6	22.31	61	-0.01	025.43	0
025	Iron (ppm)	0123	703.5	29.00	700.3	131.6	22.31	61	0.02	025.41	0
025	Iron (ppm)	0407	704.8	64.27	700.3	131.6	22.31	61	0.03	025.43	0
025	Iron (ppm)	0100	706.0	24.00	700.3	131.6	22.31	61	0.04	025.99	0
025	Iron (ppm)	0171	708.7	5.900	700.3	131.6	22.31	61	0.06	025.41	0
025	Iron (ppm)	2196	711.2	0.0000	700.3	131.6	22.31	61	0.08	025.31	0
025	Iron (ppm)	0407	713.5	19.86	700.3	131.6	22.31	61	0.10	025.41	0
025	Iron (ppm)	0358	718.5	65.00	700.3	131.6	22.31	61	0.14	025.41	0
025	Iron (ppm)	0870	724.1	7.500	700.3	131.6	22.31	61	0.18	025.43	0
025	Iron (ppm)	0520	735.0	24.00	700.3	131.6	22.31	61	0.26	025.41	0
025	Iron (ppm)	0098	737.1	12.20	700.3	131.6	22.31	61	0.28	025.41	0
025	Iron (ppm)	0353	737.5	19.00	700.3	131.6	22.31	61	0.28	025.43	0
025	Iron (ppm)	0553	743.5	23.00	700.3	131.6	22.31	61	0.33	025.53	0
025	Iron (ppm)	0651	746.8	23.80	700.3	131.6	22.31	61	0.35	025.31	0
025	Iron (ppm)	2129	749.0	40.00	700.3	131.6	22.31	61	0.37	025.42	0
025	Iron (ppm)	0675	761.3	12.01	700.3	131.6	22.31	61	0.46	025.43	0
025	Iron (ppm)	0263	768.8	4.750	700.3	131.6	22.31	61	0.52	025.42	0
025	Iron (ppm)	0208	779.5	73.40	700.3	131.6	22.31	61	0.60	025.41	0
025	Iron (ppm)	2188	785.2	0.0000	700.3	131.6	22.31	61	0.65	025.31	0
025	Iron (ppm)	0049	791.0	8.970	700.3	131.6	22.31	61	0.69	025.41	0
025	Iron (ppm)	0563	800.6	4.907	700.3	131.6	22.31	61	0.76	025.31	0
025	Iron (ppm)	0510	804.0	6.000	700.3	131.6	22.31	61	0.79	025.43	0
025	Iron (ppm)	0035	806.5	1.000	700.3	131.6	22.31	61	0.81	025.42	0
025	Iron (ppm)	2192	808.5	31.00	700.3	131.6	22.31	61	0.82	025.43	0
025	Iron (ppm)	0939	808.5	27.00	700.3	131.6	22.31	61	0.82	025.53	0
025	Iron (ppm)	0017	813.2	48.80	700.3	131.6	22.31	61	0.86	025.43	0
025	Iron (ppm)	0682	813.4	0.0000	700.3	131.6	22.31	61	0.86	025.43	0
025	Iron (ppm)	0208	821.5	11.00	700.3	131.6	22.31	61	0.92	025.31	0
025	Iron (ppm)	0148	822.6	10.78	700.3	131.6	22.31	61	0.93	025.41	0
025	Iron (ppm)	0169	827.5	5.000	700.3	131.6	22.31	61	0.97	025.43	0
025	Iron (ppm)	0098	830.1	47.50	700.3	131.6	22.31	61	0.99	025.43	0
025	Iron (ppm)	0407	848.9	65.58	700.3	131.6	22.31	61	1.13	025.53	0
025	Iron (ppm)	0035	852.0	10.00	700.3	131.6	22.31	61	1.15	025.43	0
025	Iron (ppm)	0610	881.5	27.00	700.3	131.6	22.31	61	1.38	025.43	0
025	Iron (ppm)	0038	906.3	94.20	700.3	131.6	22.31	61	1.57	025.43	0
025	Iron (ppm)	2326	938.8	15.67	700.3	131.6	22.31	61	1.81	025.43	0
025	Iron (ppm)	0042	338.5	225.0	700.3	131.6	22.31	61	-2.75	025.42	1
027	Magnesium (%)	0620	0.1361	0.0001	0.3015	0.0170	0.0067	86	-9.74	027.31	0
027	Magnesium (%)	0407	0.2432	0.0050	0.3015	0.0170	0.0067	86	-3.43	027.43	0
027	Magnesium (%)	0407	0.2524	0.0033	0.3015	0.0170	0.0067	86	-2.89	027.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
027	Magnesium (%)	0169	0.2550	0.0100	0.3015	0.0170	0.0067	86	-2.74	027.32	0
027	Magnesium (%)	0529	0.2665	0.0030	0.3015	0.0170	0.0067	86	-2.06	027.31	0
027	Magnesium (%)	0066	0.2700	0.0060	0.3015	0.0170	0.0067	86	-1.86	027.41	0
027	Magnesium (%)	0278	0.2750	0.0100	0.3015	0.0170	0.0067	86	-1.56	027.42	0
027	Magnesium (%)	2129	0.2780	0.0060	0.3015	0.0170	0.0067	86	-1.39	027.42	0
027	Magnesium (%)	0968	0.2785	0.0010	0.3015	0.0170	0.0067	86	-1.36	027.43	0
027	Magnesium (%)	0536	0.2800	0.0000	0.3015	0.0170	0.0067	86	-1.27	027.31	0
027	Magnesium (%)	0010	0.2800	0.0000	0.3015	0.0170	0.0067	86	-1.27	027.33	0
027	Magnesium (%)	0049	0.2800	0.0000	0.3015	0.0170	0.0067	86	-1.27	027.41	0
027	Magnesium (%)	0910	0.2800	0.0000	0.3015	0.0170	0.0067	86	-1.27	027.41	0
027	Magnesium (%)	0006	0.2815	0.0150	0.3015	0.0170	0.0067	86	-1.18	027.43	0
027	Magnesium (%)	0242	0.2850	0.0100	0.3015	0.0170	0.0067	86	-0.97	027.99	0
027	Magnesium (%)	0870	0.2883	0.0023	0.3015	0.0170	0.0067	86	-0.78	027.42	0
027	Magnesium (%)	0123	0.2900	0.0000	0.3015	0.0170	0.0067	86	-0.68	027.41	0
027	Magnesium (%)	0208	0.2900	0.0100	0.3015	0.0170	0.0067	86	-0.68	027.41	0
027	Magnesium (%)	0294	0.2900	0.0000	0.3015	0.0170	0.0067	86	-0.68	027.42	0
027	Magnesium (%)	0083	0.2900	0.0000	0.3015	0.0170	0.0067	86	-0.68	027.43	0
027	Magnesium (%)	0510	0.2900	0.0000	0.3015	0.0170	0.0067	86	-0.68	027.43	0
027	Magnesium (%)	0598	0.2902	0.0002	0.3015	0.0170	0.0067	86	-0.67	027.41	0
027	Magnesium (%)	0345	0.2915	0.0010	0.3015	0.0170	0.0067	86	-0.59	027.43	0
027	Magnesium (%)	0202	0.2940	0.0060	0.3015	0.0170	0.0067	86	-0.44	027.43	0
027	Magnesium (%)	2058	0.2945	0.0050	0.3015	0.0170	0.0067	86	-0.41	027.43	0
027	Magnesium (%)	0004	0.2950	0.0100	0.3015	0.0170	0.0067	86	-0.39	027.41	0
027	Magnesium (%)	0265	0.2950	0.0100	0.3015	0.0170	0.0067	86	-0.39	027.42	0
027	Magnesium (%)	0366	0.2950	0.0100	0.3015	0.0170	0.0067	86	-0.39	027.42	0
027	Magnesium (%)	0726	0.2950	0.0100	0.3015	0.0170	0.0067	86	-0.39	027.42	0
027	Magnesium (%)	0505	0.2950	0.0100	0.3015	0.0170	0.0067	86	-0.39	027.43	0
027	Magnesium (%)	0511	0.2950	0.0300	0.3015	0.0170	0.0067	86	-0.39	027.43	0
027	Magnesium (%)	0009	0.2955	0.0014	0.3015	0.0170	0.0067	86	-0.36	027.42	0
027	Magnesium (%)	0142	0.2956	0.0018	0.3015	0.0170	0.0067	86	-0.35	027.31	0
027	Magnesium (%)	0154	0.2957	0.0169	0.3015	0.0170	0.0067	86	-0.35	027.52	0
027	Magnesium (%)	0226	0.2968	0.0056	0.3015	0.0170	0.0067	86	-0.28	027.41	0
027	Magnesium (%)	0263	0.2970	0.0000	0.3015	0.0170	0.0067	86	-0.27	027.42	0
027	Magnesium (%)	0553	0.2990	0.0120	0.3015	0.0170	0.0067	86	-0.15	027.53	0
027	Magnesium (%)	2196	0.3000	0.0000	0.3015	0.0170	0.0067	86	-0.09	027.31	0
027	Magnesium (%)	2319	0.3000	0.0000	0.3015	0.0170	0.0067	86	-0.09	027.42	0
027	Magnesium (%)	2192	0.3000	0.0000	0.3015	0.0170	0.0067	86	-0.09	027.43	0
027	Magnesium (%)	0047	0.3000	0.0200	0.3015	0.0170	0.0067	86	-0.09	027.52	0
027	Magnesium (%)	0017	0.3009	0.0152	0.3015	0.0170	0.0067	86	-0.04	027.43	0
027	Magnesium (%)	0033	0.3030	0.0080	0.3015	0.0170	0.0067	86	0.09	027.43	0
027	Magnesium (%)	2322	0.3035	0.0090	0.3015	0.0170	0.0067	86	0.12	027.43	0
027	Magnesium (%)	0019	0.3050	0.0100	0.3015	0.0170	0.0067	86	0.20	027.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
027	Magnesium (%)	0098	0.3050	0.0100	0.3015	0.0170	0.0067	86	0.20	027.41	0
027	Magnesium (%)	0171	0.3050	0.0020	0.3015	0.0170	0.0067	86	0.20	027.41	0
027	Magnesium (%)	0045	0.3050	0.0040	0.3015	0.0170	0.0067	86	0.20	027.42	0
027	Magnesium (%)	0675	0.3050	0.0100	0.3015	0.0170	0.0067	86	0.20	027.43	0
027	Magnesium (%)	0914	0.3050	0.0300	0.3015	0.0170	0.0067	86	0.20	027.43	0
027	Magnesium (%)	2089	0.3050	0.0100	0.3015	0.0170	0.0067	86	0.20	027.43	0
027	Magnesium (%)	0939	0.3050	0.0100	0.3015	0.0170	0.0067	86	0.20	027.53	0
027	Magnesium (%)	0027	0.3057	0.0033	0.3015	0.0170	0.0067	86	0.24	027.43	0
027	Magnesium (%)	0001	0.3063	0.0022	0.3015	0.0170	0.0067	86	0.28	027.31	0
027	Magnesium (%)	0014	0.3070	0.0060	0.3015	0.0170	0.0067	86	0.32	027.31	0
027	Magnesium (%)	0357	0.3075	0.0040	0.3015	0.0170	0.0067	86	0.35	027.42	0
027	Magnesium (%)	0037	0.3080	0.0100	0.3015	0.0170	0.0067	86	0.38	027.42	0
027	Magnesium (%)	0038	0.3090	0.0260	0.3015	0.0170	0.0067	86	0.44	027.43	0
027	Magnesium (%)	0619	0.3095	0.0010	0.3015	0.0170	0.0067	86	0.47	027.41	0
027	Magnesium (%)	0358	0.3100	0.0200	0.3015	0.0170	0.0067	86	0.50	027.41	0
027	Magnesium (%)	0674	0.3100	0.0000	0.3015	0.0170	0.0067	86	0.50	027.41	0
027	Magnesium (%)	0720	0.3100	0.0000	0.3015	0.0170	0.0067	86	0.50	027.41	0
027	Magnesium (%)	0035	0.3100	0.0000	0.3015	0.0170	0.0067	86	0.50	027.42	0
027	Magnesium (%)	0692	0.3100	0.0000	0.3015	0.0170	0.0067	86	0.50	027.42	0
027	Magnesium (%)	0870	0.3104	0.0177	0.3015	0.0170	0.0067	86	0.52	027.43	0
027	Magnesium (%)	0425	0.3117	0.0059	0.3015	0.0170	0.0067	86	0.59	027.43	0
027	Magnesium (%)	0208	0.3125	0.0050	0.3015	0.0170	0.0067	86	0.64	027.31	0
027	Magnesium (%)	0563	0.3144	0.0005	0.3015	0.0170	0.0067	86	0.75	027.31	0
027	Magnesium (%)	0354	0.3150	0.0100	0.3015	0.0170	0.0067	86	0.79	027.41	0
027	Magnesium (%)	0098	0.3150	0.0100	0.3015	0.0170	0.0067	86	0.79	027.43	0
027	Magnesium (%)	0610	0.3150	0.0100	0.3015	0.0170	0.0067	86	0.79	027.43	0
027	Magnesium (%)	2342	0.3150	0.0100	0.3015	0.0170	0.0067	86	0.79	027.44	0
027	Magnesium (%)	0148	0.3190	0.0040	0.3015	0.0170	0.0067	86	1.03	027.41	0
027	Magnesium (%)	0190	0.3200	0.0000	0.3015	0.0170	0.0067	86	1.09	027.42	0
027	Magnesium (%)	0035	0.3200	0.0000	0.3015	0.0170	0.0067	86	1.09	027.43	0
027	Magnesium (%)	0682	0.3200	0.0000	0.3015	0.0170	0.0067	86	1.09	027.43	0
027	Magnesium (%)	0100	0.3200	0.0000	0.3015	0.0170	0.0067	86	1.09	027.99	0
027	Magnesium (%)	0353	0.3250	0.0100	0.3015	0.0170	0.0067	86	1.38	027.43	0
027	Magnesium (%)	0572	0.3290	0.0040	0.3015	0.0170	0.0067	86	1.62	027.53	0
027	Magnesium (%)	0520	0.3300	0.0200	0.3015	0.0170	0.0067	86	1.67	027.41	0
027	Magnesium (%)	0693	0.3310	0.0100	0.3015	0.0170	0.0067	86	1.73	027.42	0
027	Magnesium (%)	0042	0.3320	0.0060	0.3015	0.0170	0.0067	86	1.79	027.42	0
027	Magnesium (%)	0011	0.3500	0.0200	0.3015	0.0170	0.0067	86	2.85	027.41	0
027	Magnesium (%)	2326	0.3500	0.0000	0.3015	0.0170	0.0067	86	2.85	027.43	0
027	Magnesium (%)	0407	0.3694	0.0150	0.3015	0.0170	0.0067	86	3.99	027.53	0
027	Magnesium (%)	2302	0.4400	0.0000	0.3015	0.0170	0.0067	86	8.14	027.99	0
027	Magnesium (%)	0297	0.8150	0.0500	0.3015	0.0170	0.0067	86	30.20	027.43	2

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
028	Manganese (ppm)	0689	143.2	1.300	227.3	19.52	6.327	83	-4.31	028.31	0
028	Manganese (ppm)	0720	173.9	2.100	227.3	19.52	6.327	83	-2.74	028.41	0
028	Manganese (ppm)	0047	184.6	9.390	227.3	19.52	6.327	83	-2.19	028.52	0
028	Manganese (ppm)	0536	193.9	1.890	227.3	19.52	6.327	83	-1.71	028.31	0
028	Manganese (ppm)	0910	199.0	14.00	227.3	19.52	6.327	83	-1.45	028.41	0
028	Manganese (ppm)	0354	199.1	3.850	227.3	19.52	6.327	83	-1.45	028.41	0
028	Manganese (ppm)	0049	201.0	7.650	227.3	19.52	6.327	83	-1.35	028.41	0
028	Manganese (ppm)	0511	202.0	2.000	227.3	19.52	6.327	83	-1.30	028.43	0
028	Manganese (ppm)	0066	203.0	5.687	227.3	19.52	6.327	83	-1.25	028.44	0
028	Manganese (ppm)	0123	203.5	13.00	227.3	19.52	6.327	83	-1.22	028.41	0
028	Manganese (ppm)	0019	203.9	8.200	227.3	19.52	6.327	83	-1.20	028.41	0
028	Manganese (ppm)	0529	204.9	2.200	227.3	19.52	6.327	83	-1.15	028.31	0
028	Manganese (ppm)	0520	205.5	5.000	227.3	19.52	6.327	83	-1.12	028.41	0
028	Manganese (ppm)	0619	205.5	3.000	227.3	19.52	6.327	83	-1.12	028.41	0
028	Manganese (ppm)	0242	210.0	6.000	227.3	19.52	6.327	83	-0.89	028.99	0
028	Manganese (ppm)	0598	210.5	1.278	227.3	19.52	6.327	83	-0.86	028.41	0
028	Manganese (ppm)	0171	211.2	2.400	227.3	19.52	6.327	83	-0.83	028.41	0
028	Manganese (ppm)	0407	212.8	4.106	227.3	19.52	6.327	83	-0.74	028.43	0
028	Manganese (ppm)	0263	213.7	2.060	227.3	19.52	6.327	83	-0.70	028.42	0
028	Manganese (ppm)	0921	214.2	14.99	227.3	19.52	6.327	83	-0.67	028.31	0
028	Manganese (ppm)	0098	214.7	0.7000	227.3	19.52	6.327	83	-0.65	028.41	0
028	Manganese (ppm)	0884	215.8	2.100	227.3	19.52	6.327	83	-0.59	028.31	0
028	Manganese (ppm)	0425	216.5	0.7000	227.3	19.52	6.327	83	-0.56	028.43	0
028	Manganese (ppm)	0208	217.0	10.30	227.3	19.52	6.327	83	-0.53	028.41	0
028	Manganese (ppm)	0651	217.3	25.50	227.3	19.52	6.327	83	-0.51	028.31	0
028	Manganese (ppm)	2129	217.5	5.000	227.3	19.52	6.327	83	-0.50	028.42	0
028	Manganese (ppm)	0870	217.7	4.600	227.3	19.52	6.327	83	-0.49	028.42	0
028	Manganese (ppm)	0407	218.0	4.622	227.3	19.52	6.327	83	-0.48	028.41	0
028	Manganese (ppm)	2322	218.7	0.2000	227.3	19.52	6.327	83	-0.44	028.43	0
028	Manganese (ppm)	0004	218.8	5.190	227.3	19.52	6.327	83	-0.44	028.41	0
028	Manganese (ppm)	0038	219.6	9.500	227.3	19.52	6.327	83	-0.40	028.32	0
028	Manganese (ppm)	0553	220.0	4.000	227.3	19.52	6.327	83	-0.38	028.53	0
028	Manganese (ppm)	0968	220.1	0.1640	227.3	19.52	6.327	83	-0.37	028.43	0
028	Manganese (ppm)	0226	220.6	4.050	227.3	19.52	6.327	83	-0.34	028.41	0
028	Manganese (ppm)	0505	223.7	2.500	227.3	19.52	6.327	83	-0.19	028.43	0
028	Manganese (ppm)	0001	225.4	13.30	227.3	19.52	6.327	83	-0.10	028.31	0
028	Manganese (ppm)	2196	225.5	0.0000	227.3	19.52	6.327	83	-0.09	028.31	0
028	Manganese (ppm)	0009	226.5	11.00	227.3	19.52	6.327	83	-0.04	028.42	0
028	Manganese (ppm)	0208	227.0	6.000	227.3	19.52	6.327	83	-0.02	028.31	0
028	Manganese (ppm)	0870	227.3	13.30	227.3	19.52	6.327	83	0.00	028.43	0
028	Manganese (ppm)	0358	227.6	9.700	227.3	19.52	6.327	83	0.01	028.41	0
028	Manganese (ppm)	0563	228.9	5.371	227.3	19.52	6.327	83	0.08	028.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
028	Manganese (ppm)	2188	229.3	0.0000	227.3	19.52	6.327	83	0.10	028.31	0
028	Manganese (ppm)	0033	229.5	9.000	227.3	19.52	6.327	83	0.11	028.43	0
028	Manganese (ppm)	0169	229.5	1.000	227.3	19.52	6.327	83	0.11	028.43	0
028	Manganese (ppm)	0345	230.0	6.000	227.3	19.52	6.327	83	0.14	028.43	0
028	Manganese (ppm)	0045	230.5	9.000	227.3	19.52	6.327	83	0.16	028.42	0
028	Manganese (ppm)	2319	230.5	9.000	227.3	19.52	6.327	83	0.16	028.42	0
028	Manganese (ppm)	2058	230.8	1.700	227.3	19.52	6.327	83	0.18	028.43	0
028	Manganese (ppm)	2089	232.3	0.9900	227.3	19.52	6.327	83	0.26	028.43	0
028	Manganese (ppm)	0160	233.0	2.000	227.3	19.52	6.327	83	0.29	028.44	0
028	Manganese (ppm)	0278	233.5	3.000	227.3	19.52	6.327	83	0.32	028.42	0
028	Manganese (ppm)	0510	234.0	8.000	227.3	19.52	6.327	83	0.34	028.43	0
028	Manganese (ppm)	0693	234.3	23.42	227.3	19.52	6.327	83	0.36	028.42	0
028	Manganese (ppm)	0265	234.5	5.000	227.3	19.52	6.327	83	0.37	028.42	0
028	Manganese (ppm)	0939	235.5	5.000	227.3	19.52	6.327	83	0.42	028.53	0
028	Manganese (ppm)	0148	237.1	1.146	227.3	19.52	6.327	83	0.50	028.41	0
028	Manganese (ppm)	0042	237.5	1.000	227.3	19.52	6.327	83	0.52	028.42	0
028	Manganese (ppm)	0098	237.9	3.900	227.3	19.52	6.327	83	0.54	028.43	0
028	Manganese (ppm)	0017	238.5	13.89	227.3	19.52	6.327	83	0.57	028.43	0
028	Manganese (ppm)	0914	239.7	9.400	227.3	19.52	6.327	83	0.63	028.43	0
028	Manganese (ppm)	0610	240.0	8.000	227.3	19.52	6.327	83	0.65	028.43	0
028	Manganese (ppm)	0572	240.0	2.000	227.3	19.52	6.327	83	0.65	028.53	0
028	Manganese (ppm)	0674	240.7	0.0000	227.3	19.52	6.327	83	0.69	028.41	0
028	Manganese (ppm)	0035	241.0	6.000	227.3	19.52	6.327	83	0.70	028.42	0
028	Manganese (ppm)	0353	241.5	15.00	227.3	19.52	6.327	83	0.73	028.43	0
028	Manganese (ppm)	0357	242.8	1.983	227.3	19.52	6.327	83	0.79	028.42	0
028	Manganese (ppm)	0366	243.5	15.00	227.3	19.52	6.327	83	0.83	028.42	0
028	Manganese (ppm)	0100	244.0	2.000	227.3	19.52	6.327	83	0.85	028.99	0
028	Manganese (ppm)	0083	244.5	7.000	227.3	19.52	6.327	83	0.88	028.43	0
028	Manganese (ppm)	0675	245.7	10.61	227.3	19.52	6.327	83	0.94	028.43	0
028	Manganese (ppm)	0692	246.9	18.10	227.3	19.52	6.327	83	1.00	028.42	0
028	Manganese (ppm)	0294	247.4	4.800	227.3	19.52	6.327	83	1.03	028.42	0
028	Manganese (ppm)	0038	252.5	18.40	227.3	19.52	6.327	83	1.29	028.43	0
028	Manganese (ppm)	0726	253.0	13.50	227.3	19.52	6.327	83	1.31	028.42	0
028	Manganese (ppm)	2192	255.0	6.000	227.3	19.52	6.327	83	1.42	028.53	0
028	Manganese (ppm)	0035	256.5	1.000	227.3	19.52	6.327	83	1.49	028.43	0
028	Manganese (ppm)	0027	257.0	12.55	227.3	19.52	6.327	83	1.52	028.43	0
028	Manganese (ppm)	2326	257.6	0.1000	227.3	19.52	6.327	83	1.55	028.43	0
028	Manganese (ppm)	0037	259.0	20.90	227.3	19.52	6.327	83	1.62	028.42	0
028	Manganese (ppm)	0190	262.4	1.100	227.3	19.52	6.327	83	1.79	028.42	0
028	Manganese (ppm)	0682	276.9	0.0000	227.3	19.52	6.327	83	2.54	028.43	0
028	Manganese (ppm)	0407	292.4	0.7217	227.3	19.52	6.327	83	3.33	028.53	0
028	Manganese (ppm)	0202	238.0	30.00	227.3	19.52	6.327	83	0.55	028.43	1

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
028	Manganese (ppm)	0297	452.0	4.000	227.3	19.52	6.327	83	11.51	028.43	2
031	Phosphorus (%)	0202	0.5905	0.0090	0.7014	0.0339	0.0159	110	-3.27	031.43	0
031	Phosphorus (%)	0278	0.6050	0.0300	0.7014	0.0339	0.0159	110	-2.84	031.42	0
031	Phosphorus (%)	2322	0.6075	0.0050	0.7014	0.0339	0.0159	110	-2.77	031.43	0
031	Phosphorus (%)	0047	0.6150	0.0700	0.7014	0.0339	0.0159	110	-2.55	031.42	0
031	Phosphorus (%)	0108	0.6200	0.0400	0.7014	0.0339	0.0159	110	-2.40	031.01	0
031	Phosphorus (%)	0242	0.6350	0.0100	0.7014	0.0339	0.0159	110	-1.96	031.99	0
031	Phosphorus (%)	0154	0.6447	0.0397	0.7014	0.0339	0.0159	110	-1.67	031.52	0
031	Phosphorus (%)	0674	0.6500	0.0000	0.7014	0.0339	0.0159	110	-1.52	031.41	0
031	Phosphorus (%)	0208	0.6535	0.0350	0.7014	0.0339	0.0159	110	-1.41	031.41	0
031	Phosphorus (%)	0407	0.6547	0.0089	0.7014	0.0339	0.0159	110	-1.38	031.43	0
031	Phosphorus (%)	0294	0.6550	0.0100	0.7014	0.0339	0.0159	110	-1.37	031.42	0
031	Phosphorus (%)	0006	0.6575	0.0410	0.7014	0.0339	0.0159	110	-1.30	031.43	0
031	Phosphorus (%)	2302	0.6600	0.0000	0.7014	0.0339	0.0159	110	-1.22	031.01	0
031	Phosphorus (%)	0692	0.6600	0.0400	0.7014	0.0339	0.0159	110	-1.22	031.42	0
031	Phosphorus (%)	2302	0.6600	0.0000	0.7014	0.0339	0.0159	110	-1.22	031.99	0
031	Phosphorus (%)	0968	0.6645	0.0010	0.7014	0.0339	0.0159	110	-1.09	031.43	0
031	Phosphorus (%)	0019	0.6650	0.0100	0.7014	0.0339	0.0159	110	-1.07	031.41	0
031	Phosphorus (%)	0910	0.6650	0.0300	0.7014	0.0339	0.0159	110	-1.07	031.41	0
031	Phosphorus (%)	0265	0.6650	0.0300	0.7014	0.0339	0.0159	110	-1.07	031.42	0
031	Phosphorus (%)	0939	0.6700	0.0000	0.7014	0.0339	0.0159	110	-0.93	031.01	0
031	Phosphorus (%)	0536	0.6700	0.0000	0.7014	0.0339	0.0159	110	-0.93	031.06	0
031	Phosphorus (%)	0848	0.6700	0.0000	0.7014	0.0339	0.0159	110	-0.93	031.41	0
031	Phosphorus (%)	0366	0.6700	0.0200	0.7014	0.0339	0.0159	110	-0.93	031.42	0
031	Phosphorus (%)	0038	0.6776	0.0007	0.7014	0.0339	0.0159	110	-0.70	031.01	0
031	Phosphorus (%)	2259	0.6795	0.0030	0.7014	0.0339	0.0159	110	-0.65	031.41	0
031	Phosphorus (%)	0870	0.6797	0.0287	0.7014	0.0339	0.0159	110	-0.64	031.42	0
031	Phosphorus (%)	0011	0.6800	0.0200	0.7014	0.0339	0.0159	110	-0.63	031.41	0
031	Phosphorus (%)	0357	0.6840	0.0076	0.7014	0.0339	0.0159	110	-0.51	031.42	0
031	Phosphorus (%)	0142	0.6846	0.0025	0.7014	0.0339	0.0159	110	-0.50	031.01	0
031	Phosphorus (%)	0263	0.6890	0.0020	0.7014	0.0339	0.0159	110	-0.37	031.42	0
031	Phosphorus (%)	0035	0.6900	0.0200	0.7014	0.0339	0.0159	110	-0.34	031.42	0
031	Phosphorus (%)	0345	0.6900	0.0060	0.7014	0.0339	0.0159	110	-0.34	031.43	0
031	Phosphorus (%)	0675	0.6900	0.0000	0.7014	0.0339	0.0159	110	-0.34	031.43	0
031	Phosphorus (%)	2058	0.6900	0.0400	0.7014	0.0339	0.0159	110	-0.34	031.43	0
031	Phosphorus (%)	0100	0.6900	0.0000	0.7014	0.0339	0.0159	110	-0.34	031.99	0
031	Phosphorus (%)	0042	0.6905	0.0310	0.7014	0.0339	0.0159	110	-0.32	031.42	0
031	Phosphorus (%)	0066	0.6915	0.0290	0.7014	0.0339	0.0159	110	-0.29	031.44	0
031	Phosphorus (%)	0881	0.6920	0.0080	0.7014	0.0339	0.0159	110	-0.28	031.01	0
031	Phosphorus (%)	0693	0.6920	0.0200	0.7014	0.0339	0.0159	110	-0.28	031.42	0
031	Phosphorus (%)	0017	0.6936	0.0036	0.7014	0.0339	0.0159	110	-0.23	031.43	0
031	Phosphorus (%)	0037	0.6945	0.0330	0.7014	0.0339	0.0159	110	-0.20	031.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	2190	0.6946	0.0001	0.7014	0.0339	0.0159	110	-0.20	031.01	0
031	Phosphorus (%)	0529	0.6950	0.0100	0.7014	0.0339	0.0159	110	-0.19	031.01	0
031	Phosphorus (%)	2325	0.6950	0.0100	0.7014	0.0339	0.0159	110	-0.19	031.01	0
031	Phosphorus (%)	0004	0.6950	0.0100	0.7014	0.0339	0.0159	110	-0.19	031.41	0
031	Phosphorus (%)	0171	0.6950	0.0100	0.7014	0.0339	0.0159	110	-0.19	031.41	0
031	Phosphorus (%)	0208	0.6955	0.0030	0.7014	0.0339	0.0159	110	-0.18	031.01	0
031	Phosphorus (%)	0651	0.6958	0.0138	0.7014	0.0339	0.0159	110	-0.17	031.01	0
031	Phosphorus (%)	0014	0.6990	0.0140	0.7014	0.0339	0.0159	110	-0.07	031.42	0
031	Phosphorus (%)	0884	0.7000	0.0200	0.7014	0.0339	0.0159	110	-0.04	031.01	0
031	Phosphorus (%)	0169	0.7000	0.0000	0.7014	0.0339	0.0159	110	-0.04	031.03	0
031	Phosphorus (%)	0049	0.7000	0.0200	0.7014	0.0339	0.0159	110	-0.04	031.41	0
031	Phosphorus (%)	0358	0.7000	0.0200	0.7014	0.0339	0.0159	110	-0.04	031.41	0
031	Phosphorus (%)	0682	0.7000	0.0000	0.7014	0.0339	0.0159	110	-0.04	031.43	0
031	Phosphorus (%)	2089	0.7000	0.0200	0.7014	0.0339	0.0159	110	-0.04	031.43	0
031	Phosphorus (%)	2129	0.7005	0.0210	0.7014	0.0339	0.0159	110	-0.03	031.42	0
031	Phosphorus (%)	2006	0.7010	0.0020	0.7014	0.0339	0.0159	110	-0.01	031.01	0
031	Phosphorus (%)	0407	0.7018	0.0195	0.7014	0.0339	0.0159	110	0.01	031.41	0
031	Phosphorus (%)	2009	0.7020	0.0002	0.7014	0.0339	0.0159	110	0.02	031.01	0
031	Phosphorus (%)	0626	0.7035	0.0130	0.7014	0.0339	0.0159	110	0.06	031.01	0
031	Phosphorus (%)	0885	0.7037	0.0106	0.7014	0.0339	0.0159	110	0.07	031.01	0
031	Phosphorus (%)	0001	0.7040	0.0120	0.7014	0.0339	0.0159	110	0.08	031.43	0
031	Phosphorus (%)	0001	0.7045	0.0001	0.7014	0.0339	0.0159	110	0.09	031.03	0
031	Phosphorus (%)	2192	0.7050	0.0100	0.7014	0.0339	0.0159	110	0.10	031.01	0
031	Phosphorus (%)	0123	0.7050	0.0100	0.7014	0.0339	0.0159	110	0.10	031.41	0
031	Phosphorus (%)	0726	0.7050	0.0100	0.7014	0.0339	0.0159	110	0.10	031.42	0
031	Phosphorus (%)	0010	0.7050	0.0100	0.7014	0.0339	0.0159	110	0.10	031.43	0
031	Phosphorus (%)	0045	0.7060	0.0580	0.7014	0.0339	0.0159	110	0.13	031.42	0
031	Phosphorus (%)	0036	0.7061	0.0112	0.7014	0.0339	0.0159	110	0.14	031.03	0
031	Phosphorus (%)	0619	0.7075	0.0210	0.7014	0.0339	0.0159	110	0.18	031.41	0
031	Phosphorus (%)	0144	0.7100	0.0000	0.7014	0.0339	0.0159	110	0.25	031.41	0
031	Phosphorus (%)	0083	0.7100	0.0000	0.7014	0.0339	0.0159	110	0.25	031.43	0
031	Phosphorus (%)	0939	0.7100	0.0200	0.7014	0.0339	0.0159	110	0.25	031.53	0
031	Phosphorus (%)	0036	0.7110	0.0017	0.7014	0.0339	0.0159	110	0.28	031.43	0
031	Phosphorus (%)	0027	0.7141	0.0118	0.7014	0.0339	0.0159	110	0.37	031.43	0
031	Phosphorus (%)	2109	0.7150	0.0200	0.7014	0.0339	0.0159	110	0.40	031.01	0
031	Phosphorus (%)	0354	0.7150	0.0100	0.7014	0.0339	0.0159	110	0.40	031.41	0
031	Phosphorus (%)	0035	0.7150	0.0100	0.7014	0.0339	0.0159	110	0.40	031.43	0
031	Phosphorus (%)	0226	0.7155	0.0205	0.7014	0.0339	0.0159	110	0.41	031.41	0
031	Phosphorus (%)	0033	0.7160	0.0080	0.7014	0.0339	0.0159	110	0.43	031.43	0
031	Phosphorus (%)	0009	0.7164	0.0070	0.7014	0.0339	0.0159	110	0.44	031.42	0
031	Phosphorus (%)	0563	0.7172	0.0017	0.7014	0.0339	0.0159	110	0.46	031.01	0
031	Phosphorus (%)	0610	0.7200	0.0200	0.7014	0.0339	0.0159	110	0.55	031.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	2342	0.7200	0.0400	0.7014	0.0339	0.0159	110	0.55	031.44	0
031	Phosphorus (%)	2340	0.7225	0.0110	0.7014	0.0339	0.0159	110	0.62	031.01	0
031	Phosphorus (%)	0687	0.7250	0.0500	0.7014	0.0339	0.0159	110	0.69	031.01	0
031	Phosphorus (%)	0689	0.7250	0.0100	0.7014	0.0339	0.0159	110	0.69	031.01	0
031	Phosphorus (%)	0720	0.7250	0.0100	0.7014	0.0339	0.0159	110	0.69	031.41	0
031	Phosphorus (%)	0353	0.7300	0.0200	0.7014	0.0339	0.0159	110	0.84	031.43	0
031	Phosphorus (%)	0505	0.7300	0.0200	0.7014	0.0339	0.0159	110	0.84	031.43	0
031	Phosphorus (%)	0018	0.7325	0.0070	0.7014	0.0339	0.0159	110	0.92	031.01	0
031	Phosphorus (%)	2337	0.7350	0.0100	0.7014	0.0339	0.0159	110	0.99	031.01	0
031	Phosphorus (%)	0914	0.7350	0.0100	0.7014	0.0339	0.0159	110	0.99	031.43	0
031	Phosphorus (%)	2196	0.7400	0.0000	0.7014	0.0339	0.0159	110	1.14	031.01	0
031	Phosphorus (%)	2319	0.7415	0.0050	0.7014	0.0339	0.0159	110	1.18	031.42	0
031	Phosphorus (%)	2326	0.7450	0.0100	0.7014	0.0339	0.0159	110	1.28	031.43	0
031	Phosphorus (%)	0598	0.7460	0.0075	0.7014	0.0339	0.0159	110	1.31	031.41	0
031	Phosphorus (%)	0870	0.7497	0.0406	0.7014	0.0339	0.0159	110	1.42	031.43	0
031	Phosphorus (%)	2188	0.7500	0.0000	0.7014	0.0339	0.0159	110	1.43	031.01	0
031	Phosphorus (%)	0510	0.7500	0.0200	0.7014	0.0339	0.0159	110	1.43	031.43	0
031	Phosphorus (%)	0148	0.7550	0.0100	0.7014	0.0339	0.0159	110	1.58	031.41	0
031	Phosphorus (%)	0572	0.7555	0.0090	0.7014	0.0339	0.0159	110	1.59	031.53	0
031	Phosphorus (%)	0646	0.7600	0.0400	0.7014	0.0339	0.0159	110	1.73	031.01	0
031	Phosphorus (%)	0190	0.7600	0.0000	0.7014	0.0339	0.0159	110	1.73	031.42	0
031	Phosphorus (%)	0425	0.7600	0.0200	0.7014	0.0339	0.0159	110	1.73	031.43	0
031	Phosphorus (%)	0098	0.7618	0.0576	0.7014	0.0339	0.0159	110	1.78	031.43	0
031	Phosphorus (%)	0098	0.7633	0.0552	0.7014	0.0339	0.0159	110	1.82	031.41	0
031	Phosphorus (%)	0520	0.7650	0.0300	0.7014	0.0339	0.0159	110	1.87	031.41	0
031	Phosphorus (%)	0038	0.7810	0.0180	0.7014	0.0339	0.0159	110	2.35	031.43	0
031	Phosphorus (%)	0407	0.8673	0.0584	0.7014	0.0339	0.0159	110	4.89	031.53	0
031	Phosphorus (%)	0511	0.7100	0.0800	0.7014	0.0339	0.0159	110	0.25	031.43	1
031	Phosphorus (%)	0553	0.7165	0.1090	0.7014	0.0339	0.0159	110	0.44	031.53	1
031	Phosphorus (%)	0297	1.570	0.0000	0.7014	0.0339	0.0159	110	25.60	031.43	2
032	Potassium (%)	2188	0.9000	0.0000	1.346	0.0961	0.0375	83	-4.64	032.08	0
032	Potassium (%)	0169	1.055	0.0300	1.346	0.0961	0.0375	83	-3.03	032.32	0
032	Potassium (%)	0720	1.070	0.0000	1.346	0.0961	0.0375	83	-2.88	032.41	0
032	Potassium (%)	0066	1.095	0.0510	1.346	0.0961	0.0375	83	-2.62	032.41	0
032	Potassium (%)	0529	1.135	0.0100	1.346	0.0961	0.0375	83	-2.20	032.31	0
032	Potassium (%)	0536	1.145	0.0100	1.346	0.0961	0.0375	83	-2.10	032.31	0
032	Potassium (%)	0049	1.145	0.0300	1.346	0.0961	0.0375	83	-2.10	032.41	0
032	Potassium (%)	0563	1.162	0.0004	1.346	0.0961	0.0375	83	-1.92	032.31	0
032	Potassium (%)	0598	1.193	0.0141	1.346	0.0961	0.0375	83	-1.60	032.41	0
032	Potassium (%)	0910	1.205	0.0700	1.346	0.0961	0.0375	83	-1.47	032.41	0
032	Potassium (%)	0619	1.220	0.0200	1.346	0.0961	0.0375	83	-1.31	032.41	0
032	Potassium (%)	0004	1.225	0.0300	1.346	0.0961	0.0375	83	-1.26	032.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
032	Potassium (%)	0968	1.230	0.0010	1.346	0.0961	0.0375	83	-1.22	032.43	0
032	Potassium (%)	0123	1.245	0.0100	1.346	0.0961	0.0375	83	-1.05	032.41	0
032	Potassium (%)	0610	1.245	0.0700	1.346	0.0961	0.0375	83	-1.05	032.43	0
032	Potassium (%)	0358	1.250	0.0800	1.346	0.0961	0.0375	83	-1.00	032.41	0
032	Potassium (%)	0278	1.250	0.0800	1.346	0.0961	0.0375	83	-1.00	032.42	0
032	Potassium (%)	0047	1.260	0.1400	1.346	0.0961	0.0375	83	-0.90	032.42	0
032	Potassium (%)	0019	1.265	0.0300	1.346	0.0961	0.0375	83	-0.85	032.41	0
032	Potassium (%)	0208	1.273	0.0800	1.346	0.0961	0.0375	83	-0.76	032.41	0
032	Potassium (%)	0142	1.276	0.0101	1.346	0.0961	0.0375	83	-0.73	032.31	0
032	Potassium (%)	0354	1.280	0.0400	1.346	0.0961	0.0375	83	-0.69	032.41	0
032	Potassium (%)	0242	1.285	0.0300	1.346	0.0961	0.0375	83	-0.64	032.99	0
032	Potassium (%)	2302	1.290	0.0000	1.346	0.0961	0.0375	83	-0.59	032.99	0
032	Potassium (%)	2322	1.296	0.0030	1.346	0.0961	0.0375	83	-0.53	032.43	0
032	Potassium (%)	0505	1.300	0.0200	1.346	0.0961	0.0375	83	-0.48	032.43	0
032	Potassium (%)	0294	1.305	0.0100	1.346	0.0961	0.0375	83	-0.43	032.42	0
032	Potassium (%)	0154	1.316	0.1203	1.346	0.0961	0.0375	83	-0.32	032.52	0
032	Potassium (%)	0265	1.330	0.0200	1.346	0.0961	0.0375	83	-0.17	032.42	0
032	Potassium (%)	0553	1.330	0.0000	1.346	0.0961	0.0375	83	-0.17	032.53	0
032	Potassium (%)	0692	1.335	0.0500	1.346	0.0961	0.0375	83	-0.12	032.42	0
032	Potassium (%)	2129	1.341	0.0450	1.346	0.0961	0.0375	83	-0.06	032.42	0
032	Potassium (%)	0345	1.341	0.0410	1.346	0.0961	0.0375	83	-0.06	032.43	0
032	Potassium (%)	0870	1.343	0.0210	1.346	0.0961	0.0375	83	-0.04	032.42	0
032	Potassium (%)	0148	1.345	0.0500	1.346	0.0961	0.0375	83	-0.01	032.41	0
032	Potassium (%)	0171	1.348	0.0150	1.346	0.0961	0.0375	83	0.01	032.41	0
032	Potassium (%)	0226	1.350	0.0400	1.346	0.0961	0.0375	83	0.04	032.41	0
032	Potassium (%)	0144	1.355	0.0300	1.346	0.0961	0.0375	83	0.09	032.41	0
032	Potassium (%)	2058	1.357	0.0670	1.346	0.0961	0.0375	83	0.11	032.43	0
032	Potassium (%)	0263	1.360	0.0120	1.346	0.0961	0.0375	83	0.14	032.42	0
032	Potassium (%)	0297	1.360	0.0200	1.346	0.0961	0.0375	83	0.14	032.43	0
032	Potassium (%)	2319	1.365	0.0300	1.346	0.0961	0.0375	83	0.19	032.52	0
032	Potassium (%)	0208	1.370	0.0600	1.346	0.0961	0.0375	83	0.25	032.31	0
032	Potassium (%)	0682	1.370	0.0000	1.346	0.0961	0.0375	83	0.25	032.43	0
032	Potassium (%)	0100	1.370	0.0400	1.346	0.0961	0.0375	83	0.25	032.99	0
032	Potassium (%)	0017	1.373	0.0779	1.346	0.0961	0.0375	83	0.27	032.43	0
032	Potassium (%)	0914	1.375	0.0100	1.346	0.0961	0.0375	83	0.30	032.43	0
032	Potassium (%)	0027	1.379	0.0065	1.346	0.0961	0.0375	83	0.34	032.43	0
032	Potassium (%)	0035	1.380	0.0000	1.346	0.0961	0.0375	83	0.35	032.43	0
032	Potassium (%)	0407	1.385	0.0865	1.346	0.0961	0.0375	83	0.40	032.43	0
032	Potassium (%)	0425	1.385	0.0100	1.346	0.0961	0.0375	83	0.40	032.43	0
032	Potassium (%)	0357	1.386	0.0108	1.346	0.0961	0.0375	83	0.41	032.42	0
032	Potassium (%)	0001	1.387	0.0550	1.346	0.0961	0.0375	83	0.42	032.99	0
032	Potassium (%)	0035	1.390	0.0000	1.346	0.0961	0.0375	83	0.45	032.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
032	Potassium (%)	0083	1.390	0.0000	1.346	0.0961	0.0375	83	0.45	032.43	0
032	Potassium (%)	0033	1.395	0.0100	1.346	0.0961	0.0375	83	0.51	032.43	0
032	Potassium (%)	0510	1.395	0.0100	1.346	0.0961	0.0375	83	0.51	032.43	0
032	Potassium (%)	0675	1.395	0.0100	1.346	0.0961	0.0375	83	0.51	032.43	0
032	Potassium (%)	0407	1.397	0.0076	1.346	0.0961	0.0375	83	0.53	032.41	0
032	Potassium (%)	0038	1.399	0.0560	1.346	0.0961	0.0375	83	0.55	032.43	0
032	Potassium (%)	0045	1.400	0.0800	1.346	0.0961	0.0375	83	0.56	032.42	0
032	Potassium (%)	0366	1.400	0.0800	1.346	0.0961	0.0375	83	0.56	032.42	0
032	Potassium (%)	2089	1.405	0.0100	1.346	0.0961	0.0375	83	0.61	032.43	0
032	Potassium (%)	0098	1.410	0.0200	1.346	0.0961	0.0375	83	0.66	032.41	0
032	Potassium (%)	0098	1.415	0.0100	1.346	0.0961	0.0375	83	0.71	032.43	0
032	Potassium (%)	0353	1.415	0.0500	1.346	0.0961	0.0375	83	0.71	032.43	0
032	Potassium (%)	0693	1.416	0.0280	1.346	0.0961	0.0375	83	0.72	032.42	0
032	Potassium (%)	0010	1.420	0.0400	1.346	0.0961	0.0375	83	0.77	032.43	0
032	Potassium (%)	2342	1.420	0.0600	1.346	0.0961	0.0375	83	0.77	032.44	0
032	Potassium (%)	0870	1.428	0.0800	1.346	0.0961	0.0375	83	0.85	032.43	0
032	Potassium (%)	0042	1.430	0.0600	1.346	0.0961	0.0375	83	0.87	032.42	0
032	Potassium (%)	0726	1.435	0.0300	1.346	0.0961	0.0375	83	0.92	032.42	0
032	Potassium (%)	0572	1.435	0.0500	1.346	0.0961	0.0375	83	0.92	032.53	0
032	Potassium (%)	0036	1.440	0.0333	1.346	0.0961	0.0375	83	0.97	032.43	0
032	Potassium (%)	0190	1.450	0.0200	1.346	0.0961	0.0375	83	1.08	032.42	0
032	Potassium (%)	0939	1.450	0.0600	1.346	0.0961	0.0375	83	1.08	032.53	0
032	Potassium (%)	0037	1.459	0.0780	1.346	0.0961	0.0375	83	1.17	032.42	0
032	Potassium (%)	0202	1.460	0.0000	1.346	0.0961	0.0375	83	1.18	032.43	0
032	Potassium (%)	0009	1.485	0.1005	1.346	0.0961	0.0375	83	1.44	032.42	0
032	Potassium (%)	0520	1.505	0.1100	1.346	0.0961	0.0375	83	1.65	032.41	0
032	Potassium (%)	0011	1.525	0.0700	1.346	0.0961	0.0375	83	1.86	032.41	0
032	Potassium (%)	2326	1.550	0.1000	1.346	0.0961	0.0375	83	2.12	032.43	0
032	Potassium (%)	0407	1.696	0.0636	1.346	0.0961	0.0375	83	3.64	032.53	0
032	Potassium (%)	0038	1.218	0.2040	1.346	0.0961	0.0375	83	-1.34	032.31	1
032	Potassium (%)	0511	1.320	0.2000	1.346	0.0961	0.0375	83	-0.27	032.43	1
033	Salt as chloride (%)	0536	0.7600	0.0000	1.049	0.0796	0.0232	59	-3.63	033.99	0
033	Salt as chloride (%)	0358	0.8100	0.0400	1.049	0.0796	0.0232	59	-3.00	033.99	0
033	Salt as chloride (%)	0017	0.8793	0.0500	1.049	0.0796	0.0232	59	-2.13	033.99	0
033	Salt as chloride (%)	0171	0.8900	0.0000	1.049	0.0796	0.0232	59	-2.00	033.05	0
033	Salt as chloride (%)	0968	0.9200	0.0000	1.049	0.0796	0.0232	59	-1.62	033.00	0
033	Salt as chloride (%)	0265	0.9200	0.0200	1.049	0.0796	0.0232	59	-1.62	033.03	0
033	Salt as chloride (%)	0968	0.9200	0.0000	1.049	0.0796	0.0232	59	-1.62	033.03	0
033	Salt as chloride (%)	2188	0.9200	0.0000	1.049	0.0796	0.0232	59	-1.62	033.05	0
033	Salt as chloride (%)	0038	0.9326	0.1022	1.049	0.0796	0.0232	59	-1.46	033.01	0
033	Salt as chloride (%)	0190	0.9400	0.0000	1.049	0.0796	0.0232	59	-1.37	033.03	0
033	Salt as chloride (%)	0505	0.9600	0.0800	1.049	0.0796	0.0232	59	-1.12	033.03	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
033	Salt as chloride (%)	2193	0.9750	0.0100	1.049	0.0796	0.0232	59	-0.93	033.00	0
033	Salt as chloride (%)	2009	0.9824	0.0003	1.049	0.0796	0.0232	59	-0.84	033.00	0
033	Salt as chloride (%)	0726	0.9900	0.2200	1.049	0.0796	0.0232	59	-0.74	033.03	0
033	Salt as chloride (%)	0511	0.9950	0.1700	1.049	0.0796	0.0232	59	-0.68	033.00	0
033	Salt as chloride (%)	0675	1.015	0.0100	1.049	0.0796	0.0232	59	-0.43	033.00	0
033	Salt as chloride (%)	0885	1.019	0.0066	1.049	0.0796	0.0232	59	-0.37	033.00	0
033	Salt as chloride (%)	0425	1.020	0.0200	1.049	0.0796	0.0232	59	-0.37	033.01	0
033	Salt as chloride (%)	2073	1.025	0.0100	1.049	0.0796	0.0232	59	-0.30	033.99	0
033	Salt as chloride (%)	0123	1.035	0.0300	1.049	0.0796	0.0232	59	-0.18	033.00	0
033	Salt as chloride (%)	0366	1.035	0.1300	1.049	0.0796	0.0232	59	-0.18	033.00	0
033	Salt as chloride (%)	0689	1.040	0.0000	1.049	0.0796	0.0232	59	-0.11	033.05	0
033	Salt as chloride (%)	0910	1.040	0.0400	1.049	0.0796	0.0232	59	-0.11	033.99	0
033	Salt as chloride (%)	0353	1.045	0.0100	1.049	0.0796	0.0232	59	-0.05	033.00	0
033	Salt as chloride (%)	0098	1.050	0.0000	1.049	0.0796	0.0232	59	0.01	033.01	0
033	Salt as chloride (%)	0309	1.057	0.0053	1.049	0.0796	0.0232	59	0.10	033.00	0
033	Salt as chloride (%)	0881	1.064	0.0030	1.049	0.0796	0.0232	59	0.18	033.00	0
033	Salt as chloride (%)	0004	1.065	0.0100	1.049	0.0796	0.0232	59	0.20	033.01	0
033	Salt as chloride (%)	0939	1.065	0.0100	1.049	0.0796	0.0232	59	0.20	033.01	0
033	Salt as chloride (%)	2196	1.070	0.0000	1.049	0.0796	0.0232	59	0.26	033.00	0
033	Salt as chloride (%)	0028	1.070	0.0000	1.049	0.0796	0.0232	59	0.26	033.01	0
033	Salt as chloride (%)	0083	1.070	0.0000	1.049	0.0796	0.0232	59	0.26	033.01	0
033	Salt as chloride (%)	0884	1.075	0.0100	1.049	0.0796	0.0232	59	0.33	033.00	0
033	Salt as chloride (%)	0208	1.075	0.0100	1.049	0.0796	0.0232	59	0.33	033.99	0
033	Salt as chloride (%)	0242	1.080	0.0200	1.049	0.0796	0.0232	59	0.39	033.01	0
033	Salt as chloride (%)	0948	1.080	0.0000	1.049	0.0796	0.0232	59	0.39	033.01	0
033	Salt as chloride (%)	0001	1.090	0.0200	1.049	0.0796	0.0232	59	0.51	033.01	0
033	Salt as chloride (%)	0010	1.090	0.0000	1.049	0.0796	0.0232	59	0.51	033.01	0
033	Salt as chloride (%)	0019	1.090	0.0000	1.049	0.0796	0.0232	59	0.51	033.01	0
033	Salt as chloride (%)	0100	1.090	0.0000	1.049	0.0796	0.0232	59	0.51	033.01	0
033	Salt as chloride (%)	0148	1.090	0.0000	1.049	0.0796	0.0232	59	0.51	033.01	0
033	Salt as chloride (%)	0226	1.090	0.0200	1.049	0.0796	0.0232	59	0.51	033.01	0
033	Salt as chloride (%)	0682	1.090	0.0000	1.049	0.0796	0.0232	59	0.51	033.01	0
033	Salt as chloride (%)	2319	1.090	0.0000	1.049	0.0796	0.0232	59	0.51	033.01	0
033	Salt as chloride (%)	0407	1.091	0.0080	1.049	0.0796	0.0232	59	0.53	033.01	0
033	Salt as chloride (%)	0354	1.091	0.0020	1.049	0.0796	0.0232	59	0.53	033.01	0
033	Salt as chloride (%)	2109	1.092	0.0150	1.049	0.0796	0.0232	59	0.53	033.01	0
033	Salt as chloride (%)	0045	1.095	0.0100	1.049	0.0796	0.0232	59	0.58	033.00	0
033	Salt as chloride (%)	0693	1.097	0.0170	1.049	0.0796	0.0232	59	0.60	033.00	0
033	Salt as chloride (%)	2190	1.110	0.0166	1.049	0.0796	0.0232	59	0.77	033.00	0
033	Salt as chloride (%)	2326	1.110	0.0200	1.049	0.0796	0.0232	59	0.77	033.99	0
033	Salt as chloride (%)	0278	1.120	0.0200	1.049	0.0796	0.0232	59	0.89	033.01	0
033	Salt as chloride (%)	2268	1.120	0.0000	1.049	0.0796	0.0232	59	0.89	033.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
033	Salt as chloride (%)	0539	1.125	0.0100	1.049	0.0796	0.0232	59	0.95	033.00	0
033	Salt as chloride (%)	0510	1.145	0.0100	1.049	0.0796	0.0232	59	1.21	033.01	0
033	Salt as chloride (%)	0003	1.200	0.0000	1.049	0.0796	0.0232	59	1.90	033.01	0
033	Salt as chloride (%)	2076	1.210	0.0640	1.049	0.0796	0.0232	59	2.02	033.00	0
033	Salt as chloride (%)	0619	1.210	0.0000	1.049	0.0796	0.0232	59	2.02	033.01	0
033	Salt as chloride (%)	2043	1.210	0.1200	1.049	0.0796	0.0232	59	2.02	033.05	0
033	Salt as chloride (%)	0144	1.065	0.4700	1.049	0.0796	0.0232	59	0.20	033.03	1
034	Selenium, Total (Se) (ppm)	0619	0.5165	0.0150	0.8117	0.2080	0.0639	21	-1.42	034.41	0
034	Selenium, Total (Se) (ppm)	0047	0.5750	0.0300	0.8117	0.2080	0.0639	21	-1.14	034.52	0
034	Selenium, Total (Se) (ppm)	0045	0.5855	0.0090	0.8117	0.2080	0.0639	21	-1.09	034.04	0
034	Selenium, Total (Se) (ppm)	0563	0.6579	0.0583	0.8117	0.2080	0.0639	21	-0.74	034.04	0
034	Selenium, Total (Se) (ppm)	0171	0.6585	0.0970	0.8117	0.2080	0.0639	21	-0.74	034.04	0
034	Selenium, Total (Se) (ppm)	0148	0.7290	0.0760	0.8117	0.2080	0.0639	21	-0.40	034.41	0
034	Selenium, Total (Se) (ppm)	0910	0.7450	0.0700	0.8117	0.2080	0.0639	21	-0.32	034.52	0
034	Selenium, Total (Se) (ppm)	0553	0.7470	0.0800	0.8117	0.2080	0.0639	21	-0.31	034.53	0
034	Selenium, Total (Se) (ppm)	0009	0.7490	0.1519	0.8117	0.2080	0.0639	21	-0.30	034.53	0
034	Selenium, Total (Se) (ppm)	0560	0.7660	0.0340	0.8117	0.2080	0.0639	21	-0.22	034.53	0
034	Selenium, Total (Se) (ppm)	0033	0.7765	0.0130	0.8117	0.2080	0.0639	21	-0.17	034.53	0
034	Selenium, Total (Se) (ppm)	0169	0.7800	0.0200	0.8117	0.2080	0.0639	21	-0.15	034.04	0
034	Selenium, Total (Se) (ppm)	0610	0.7800	0.0600	0.8117	0.2080	0.0639	21	-0.15	034.53	0
034	Selenium, Total (Se) (ppm)	0038	0.8275	0.0370	0.8117	0.2080	0.0639	21	0.08	034.01	0
034	Selenium, Total (Se) (ppm)	0208	0.8755	0.1210	0.8117	0.2080	0.0639	21	0.31	034.52	0
034	Selenium, Total (Se) (ppm)	0003	0.8825	0.0470	0.8117	0.2080	0.0639	21	0.34	034.53	0
034	Selenium, Total (Se) (ppm)	0010	0.9000	0.0000	0.8117	0.2080	0.0639	21	0.42	034.53	0
034	Selenium, Total (Se) (ppm)	0407	1.162	0.0172	0.8117	0.2080	0.0639	21	1.68	034.53	0
034	Selenium, Total (Se) (ppm)	0968	1.374	0.2150	0.8117	0.2080	0.0639	21	2.70	034.43	0
034	Selenium, Total (Se) (ppm)	0190	1.395	0.0700	0.8117	0.2080	0.0639	21	2.80	034.99	0
034	Selenium, Total (Se) (ppm)	0511	1.400	0.1200	0.8117	0.2080	0.0639	21	2.83	034.43	0
034	Selenium, Total (Se) (ppm)	0692	< 1		0.8117	0.2080	0.0639	21		034.42	5
034	Selenium, Total (Se) (ppm)	0682	< 5		0.8117	0.2080	0.0639	21		034.43	5
034	Selenium, Total (Se) (ppm)	0160	< 10		0.8117	0.2080	0.0639	21		034.99	5
035	Sodium (%)	0169	0.2750	0.0100	0.3532	0.0250	0.0116	83	-3.13	035.32	0
035	Sodium (%)	0598	0.2993	0.0003	0.3532	0.0250	0.0116	83	-2.16	035.41	0
035	Sodium (%)	0049	0.3000	0.0000	0.3532	0.0250	0.0116	83	-2.13	035.41	0
035	Sodium (%)	0720	0.3050	0.0300	0.3532	0.0250	0.0116	83	-1.93	035.41	0
035	Sodium (%)	0620	0.3075	0.0013	0.3532	0.0250	0.0116	83	-1.83	035.31	0
035	Sodium (%)	0242	0.3100	0.0000	0.3532	0.0250	0.0116	83	-1.73	035.99	0
035	Sodium (%)	0505	0.3150	0.0100	0.3532	0.0250	0.0116	83	-1.53	035.43	0
035	Sodium (%)	0366	0.3185	0.0170	0.3532	0.0250	0.0116	83	-1.39	035.42	0
035	Sodium (%)	0208	0.3190	0.0140	0.3532	0.0250	0.0116	83	-1.37	035.31	0
035	Sodium (%)	0358	0.3200	0.0200	0.3532	0.0250	0.0116	83	-1.33	035.41	0
035	Sodium (%)	0910	0.3200	0.0000	0.3532	0.0250	0.0116	83	-1.33	035.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
035	Sodium (%)	0619	0.3250	0.0100	0.3532	0.0250	0.0116	83	-1.13	035.41	0
035	Sodium (%)	0278	0.3250	0.0100	0.3532	0.0250	0.0116	83	-1.13	035.42	0
035	Sodium (%)	2322	0.3255	0.0050	0.3532	0.0250	0.0116	83	-1.11	035.43	0
035	Sodium (%)	0006	0.3260	0.0160	0.3532	0.0250	0.0116	83	-1.09	035.43	0
035	Sodium (%)	0066	0.3285	0.0130	0.3532	0.0250	0.0116	83	-0.99	035.41	0
035	Sodium (%)	0208	0.3310	0.0020	0.3532	0.0250	0.0116	83	-0.89	035.41	0
035	Sodium (%)	0123	0.3350	0.0100	0.3532	0.0250	0.0116	83	-0.73	035.41	0
035	Sodium (%)	0294	0.3350	0.0100	0.3532	0.0250	0.0116	83	-0.73	035.42	0
035	Sodium (%)	0345	0.3355	0.0050	0.3532	0.0250	0.0116	83	-0.71	035.43	0
035	Sodium (%)	2129	0.3370	0.0120	0.3532	0.0250	0.0116	83	-0.65	035.42	0
035	Sodium (%)	0154	0.3375	0.0208	0.3532	0.0250	0.0116	83	-0.63	035.52	0
035	Sodium (%)	0968	0.3385	0.0030	0.3532	0.0250	0.0116	83	-0.59	035.43	0
035	Sodium (%)	0610	0.3400	0.0400	0.3532	0.0250	0.0116	83	-0.53	035.43	0
035	Sodium (%)	2058	0.3405	0.0170	0.3532	0.0250	0.0116	83	-0.51	035.43	0
035	Sodium (%)	0142	0.3410	0.0117	0.3532	0.0250	0.0116	83	-0.49	035.31	0
035	Sodium (%)	0939	0.3450	0.0100	0.3532	0.0250	0.0116	83	-0.33	035.53	0
035	Sodium (%)	0017	0.3459	0.0196	0.3532	0.0250	0.0116	83	-0.29	035.43	0
035	Sodium (%)	0529	0.3490	0.0160	0.3532	0.0250	0.0116	83	-0.17	035.31	0
035	Sodium (%)	0536	0.3500	0.0200	0.3532	0.0250	0.0116	83	-0.13	035.31	0
035	Sodium (%)	0004	0.3500	0.0200	0.3532	0.0250	0.0116	83	-0.13	035.41	0
035	Sodium (%)	0019	0.3500	0.0000	0.3532	0.0250	0.0116	83	-0.13	035.41	0
035	Sodium (%)	0148	0.3500	0.0160	0.3532	0.0250	0.0116	83	-0.13	035.41	0
035	Sodium (%)	0265	0.3500	0.0000	0.3532	0.0250	0.0116	83	-0.13	035.42	0
035	Sodium (%)	0682	0.3500	0.0000	0.3532	0.0250	0.0116	83	-0.13	035.43	0
035	Sodium (%)	0563	0.3512	0.0033	0.3532	0.0250	0.0116	83	-0.08	035.31	0
035	Sodium (%)	0407	0.3535	0.0088	0.3532	0.0250	0.0116	83	0.01	035.43	0
035	Sodium (%)	0144	0.3550	0.0100	0.3532	0.0250	0.0116	83	0.07	035.41	0
035	Sodium (%)	0692	0.3550	0.0100	0.3532	0.0250	0.0116	83	0.07	035.42	0
035	Sodium (%)	2319	0.3550	0.0080	0.3532	0.0250	0.0116	83	0.07	035.42	0
035	Sodium (%)	0914	0.3550	0.0300	0.3532	0.0250	0.0116	83	0.07	035.43	0
035	Sodium (%)	0357	0.3552	0.0024	0.3532	0.0250	0.0116	83	0.08	035.42	0
035	Sodium (%)	0263	0.3555	0.0010	0.3532	0.0250	0.0116	83	0.09	035.42	0
035	Sodium (%)	0033	0.3555	0.0030	0.3532	0.0250	0.0116	83	0.09	035.43	0
035	Sodium (%)	0553	0.3560	0.0240	0.3532	0.0250	0.0116	83	0.11	035.53	0
035	Sodium (%)	0407	0.3564	0.0022	0.3532	0.0250	0.0116	83	0.13	035.41	0
035	Sodium (%)	0045	0.3575	0.0090	0.3532	0.0250	0.0116	83	0.17	035.42	0
035	Sodium (%)	0510	0.3575	0.0030	0.3532	0.0250	0.0116	83	0.17	035.43	0
035	Sodium (%)	0108	0.3600	0.0600	0.3532	0.0250	0.0116	83	0.27	035.05	0
035	Sodium (%)	2196	0.3600	0.0000	0.3532	0.0250	0.0116	83	0.27	035.31	0
035	Sodium (%)	2325	0.3600	0.0000	0.3532	0.0250	0.0116	83	0.27	035.31	0
035	Sodium (%)	0190	0.3600	0.0200	0.3532	0.0250	0.0116	83	0.27	035.42	0
035	Sodium (%)	0100	0.3600	0.0000	0.3532	0.0250	0.0116	83	0.27	035.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
035	Sodium (%)	0870	0.3607	0.0008	0.3532	0.0250	0.0116	83	0.30	035.42	0
035	Sodium (%)	0520	0.3614	0.0088	0.3532	0.0250	0.0116	83	0.33	035.41	0
035	Sodium (%)	2259	0.3630	0.0000	0.3532	0.0250	0.0116	83	0.39	035.41	0
035	Sodium (%)	2109	0.3650	0.0240	0.3532	0.0250	0.0116	83	0.47	035.01	0
035	Sodium (%)	0098	0.3650	0.0100	0.3532	0.0250	0.0116	83	0.47	035.41	0
035	Sodium (%)	0171	0.3650	0.0100	0.3532	0.0250	0.0116	83	0.47	035.41	0
035	Sodium (%)	0035	0.3650	0.0100	0.3532	0.0250	0.0116	83	0.47	035.42	0
035	Sodium (%)	0726	0.3650	0.0300	0.3532	0.0250	0.0116	83	0.47	035.42	0
035	Sodium (%)	0098	0.3650	0.0100	0.3532	0.0250	0.0116	83	0.47	035.43	0
035	Sodium (%)	0675	0.3650	0.0100	0.3532	0.0250	0.0116	83	0.47	035.43	0
035	Sodium (%)	0693	0.3665	0.0070	0.3532	0.0250	0.0116	83	0.53	035.42	0
035	Sodium (%)	0572	0.3670	0.0040	0.3532	0.0250	0.0116	83	0.55	035.53	0
035	Sodium (%)	0226	0.3681	0.0079	0.3532	0.0250	0.0116	83	0.59	035.41	0
035	Sodium (%)	0870	0.3722	0.0204	0.3532	0.0250	0.0116	83	0.76	035.43	0
035	Sodium (%)	0354	0.3750	0.0100	0.3532	0.0250	0.0116	83	0.87	035.41	0
035	Sodium (%)	2089	0.3750	0.0100	0.3532	0.0250	0.0116	83	0.87	035.43	0
035	Sodium (%)	0037	0.3775	0.0270	0.3532	0.0250	0.0116	83	0.97	035.42	0
035	Sodium (%)	0010	0.3800	0.0000	0.3532	0.0250	0.0116	83	1.07	035.43	0
035	Sodium (%)	0083	0.3800	0.0000	0.3532	0.0250	0.0116	83	1.07	035.43	0
035	Sodium (%)	0202	0.3800	0.0180	0.3532	0.0250	0.0116	83	1.07	035.43	0
035	Sodium (%)	0042	0.3820	0.0120	0.3532	0.0250	0.0116	83	1.15	035.42	0
035	Sodium (%)	0035	0.3850	0.0100	0.3532	0.0250	0.0116	83	1.27	035.43	0
035	Sodium (%)	0353	0.3859	0.0134	0.3532	0.0250	0.0116	83	1.31	035.43	0
035	Sodium (%)	0038	0.4005	0.0350	0.3532	0.0250	0.0116	83	1.89	035.43	0
035	Sodium (%)	0038	0.4063	0.0380	0.3532	0.0250	0.0116	83	2.12	035.31	0
035	Sodium (%)	2326	0.4150	0.0100	0.3532	0.0250	0.0116	83	2.47	035.43	0
035	Sodium (%)	2302	0.4200	0.0000	0.3532	0.0250	0.0116	83	2.67	035.99	0
035	Sodium (%)	0425	0.4300	0.0200	0.3532	0.0250	0.0116	83	3.07	035.43	0
035	Sodium (%)	0407	0.4310	0.0225	0.3532	0.0250	0.0116	83	3.11	035.53	0
035	Sodium (%)	2188	0.7600	0.0000	0.3532	0.0250	0.0116	83	16.26	035.01	0
035	Sodium (%)	0511	0.3350	0.0900	0.3532	0.0250	0.0116	83	-0.73	035.43	1
035	Sodium (%)	0297	1.460	0.0600	0.3532	0.0250	0.0116	83	44.25	035.43	2
036	Sulfur (%)	0720	0.0800	0.0000	0.2993	0.0298	0.0070	36	-7.35	036.04	0
036	Sulfur (%)	0171	0.2120	0.0020	0.2993	0.0298	0.0070	36	-2.93	036.42	0
036	Sulfur (%)	0407	0.2531	0.0010	0.2993	0.0298	0.0070	36	-1.55	036.43	0
036	Sulfur (%)	0242	0.2550	0.0100	0.2993	0.0298	0.0070	36	-1.49	036.99	0
036	Sulfur (%)	0278	0.2700	0.0200	0.2993	0.0298	0.0070	36	-0.98	036.42	0
036	Sulfur (%)	0870	0.2703	0.0065	0.2993	0.0298	0.0070	36	-0.97	036.42	0
036	Sulfur (%)	0366	0.2750	0.0100	0.2993	0.0298	0.0070	36	-0.82	036.42	0
036	Sulfur (%)	0692	0.2750	0.0100	0.2993	0.0298	0.0070	36	-0.82	036.42	0
036	Sulfur (%)	0294	0.2800	0.0000	0.2993	0.0298	0.0070	36	-0.65	036.42	0
036	Sulfur (%)	0553	0.2810	0.0100	0.2993	0.0298	0.0070	36	-0.61	036.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
036	Sulfur (%)	0035	0.2900	0.0000	0.2993	0.0298	0.0070	36	-0.31	036.42	0
036	Sulfur (%)	0265	0.2900	0.0200	0.2993	0.0298	0.0070	36	-0.31	036.42	0
036	Sulfur (%)	2089	0.2900	0.0200	0.2993	0.0298	0.0070	36	-0.31	036.43	0
036	Sulfur (%)	0693	0.2925	0.0050	0.2993	0.0298	0.0070	36	-0.23	036.42	0
036	Sulfur (%)	0353	0.2950	0.0100	0.2993	0.0298	0.0070	36	-0.14	036.43	0
036	Sulfur (%)	0357	0.2962	0.0004	0.2993	0.0298	0.0070	36	-0.10	036.42	0
036	Sulfur (%)	0045	0.2965	0.0010	0.2993	0.0298	0.0070	36	-0.09	036.42	0
036	Sulfur (%)	0042	0.2985	0.0110	0.2993	0.0298	0.0070	36	-0.03	036.42	0
036	Sulfur (%)	0726	0.3000	0.0000	0.2993	0.0298	0.0070	36	0.02	036.42	0
036	Sulfur (%)	2129	0.3000	0.0040	0.2993	0.0298	0.0070	36	0.02	036.42	0
036	Sulfur (%)	0345	0.3010	0.0060	0.2993	0.0298	0.0070	36	0.06	036.43	0
036	Sulfur (%)	0870	0.3025	0.0042	0.2993	0.0298	0.0070	36	0.11	036.43	0
036	Sulfur (%)	0226	0.3037	0.0011	0.2993	0.0298	0.0070	36	0.15	036.04	0
036	Sulfur (%)	0148	0.3085	0.0010	0.2993	0.0298	0.0070	36	0.31	036.04	0
036	Sulfur (%)	0035	0.3100	0.0000	0.2993	0.0298	0.0070	36	0.36	036.43	0
036	Sulfur (%)	0083	0.3100	0.0000	0.2993	0.0298	0.0070	36	0.36	036.43	0
036	Sulfur (%)	0708	0.3150	0.0020	0.2993	0.0298	0.0070	36	0.53	036.42	0
036	Sulfur (%)	0202	0.3160	0.0040	0.2993	0.0298	0.0070	36	0.56	036.43	0
036	Sulfur (%)	0098	0.3209	0.0227	0.2993	0.0298	0.0070	36	0.72	036.43	0
036	Sulfur (%)	0510	0.3300	0.0000	0.2993	0.0298	0.0070	36	1.03	036.43	0
036	Sulfur (%)	0682	0.3300	0.0000	0.2993	0.0298	0.0070	36	1.03	036.43	0
036	Sulfur (%)	2326	0.3350	0.0100	0.2993	0.0298	0.0070	36	1.20	036.43	0
036	Sulfur (%)	0038	0.3420	0.0340	0.2993	0.0298	0.0070	36	1.43	036.43	0
036	Sulfur (%)	0169	0.3550	0.0100	0.2993	0.0298	0.0070	36	1.87	036.43	0
036	Sulfur (%)	0037	0.3585	0.0150	0.2993	0.0298	0.0070	36	1.98	036.42	0
036	Sulfur (%)	0297	0.4500	0.0000	0.2993	0.0298	0.0070	36	5.05	036.43	0
036	Sulfur (%)	0407	0.3059	0.0445	0.2993	0.0298	0.0070	36	0.22	036.42	1
037	Zinc (ppm)	0914	22.58	2.060	187.3	30.14	6.178	84	-5.47	037.43	0
037	Zinc (ppm)	2188	82.55	0.0000	187.3	30.14	6.178	84	-3.48	037.31	0
037	Zinc (ppm)	0536	107.8	5.890	187.3	30.14	6.178	84	-2.64	037.31	0
037	Zinc (ppm)	0529	119.0	5.200	187.3	30.14	6.178	84	-2.27	037.31	0
037	Zinc (ppm)	0848	126.1	1.860	187.3	30.14	6.178	84	-2.03	037.41	0
037	Zinc (ppm)	0674	128.0	0.0000	187.3	30.14	6.178	84	-1.97	037.41	0
037	Zinc (ppm)	0354	131.0	8.930	187.3	30.14	6.178	84	-1.87	037.41	0
037	Zinc (ppm)	0019	132.3	8.800	187.3	30.14	6.178	84	-1.83	037.41	0
037	Zinc (ppm)	0598	132.6	2.486	187.3	30.14	6.178	84	-1.81	037.41	0
037	Zinc (ppm)	0047	133.4	10.29	187.3	30.14	6.178	84	-1.79	037.52	0
037	Zinc (ppm)	0004	138.5	4.720	187.3	30.14	6.178	84	-1.62	037.41	0
037	Zinc (ppm)	0689	140.9	0.2000	187.3	30.14	6.178	84	-1.54	037.31	0
037	Zinc (ppm)	0619	143.5	3.000	187.3	30.14	6.178	84	-1.45	037.41	0
037	Zinc (ppm)	0171	151.3	2.000	187.3	30.14	6.178	84	-1.20	037.41	0
037	Zinc (ppm)	0123	152.5	13.00	187.3	30.14	6.178	84	-1.16	037.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
037	Zinc (ppm)	2196	157.3	0.0000	187.3	30.14	6.178	84	-1.00	037.31	0
037	Zinc (ppm)	0651	158.1	7.167	187.3	30.14	6.178	84	-0.97	037.31	0
037	Zinc (ppm)	0049	159.1	26.85	187.3	30.14	6.178	84	-0.94	037.41	0
037	Zinc (ppm)	0066	161.7	2.950	187.3	30.14	6.178	84	-0.85	037.44	0
037	Zinc (ppm)	0720	166.9	1.830	187.3	30.14	6.178	84	-0.68	037.41	0
037	Zinc (ppm)	0294	168.8	1.100	187.3	30.14	6.178	84	-0.62	037.42	0
037	Zinc (ppm)	0226	170.6	1.610	187.3	30.14	6.178	84	-0.55	037.41	0
037	Zinc (ppm)	0520	171.0	4.000	187.3	30.14	6.178	84	-0.54	037.41	0
037	Zinc (ppm)	0910	171.0	14.00	187.3	30.14	6.178	84	-0.54	037.41	0
037	Zinc (ppm)	0366	174.5	11.00	187.3	30.14	6.178	84	-0.43	037.42	0
037	Zinc (ppm)	0968	177.3	2.029	187.3	30.14	6.178	84	-0.33	037.43	0
037	Zinc (ppm)	0098	177.5	2.300	187.3	30.14	6.178	84	-0.33	037.41	0
037	Zinc (ppm)	0358	178.8	10.98	187.3	30.14	6.178	84	-0.28	037.41	0
037	Zinc (ppm)	0045	180.5	3.000	187.3	30.14	6.178	84	-0.23	037.42	0
037	Zinc (ppm)	0511	181.0	2.000	187.3	30.14	6.178	84	-0.21	037.43	0
037	Zinc (ppm)	0001	181.8	3.100	187.3	30.14	6.178	84	-0.19	037.31	0
037	Zinc (ppm)	2322	183.5	1.040	187.3	30.14	6.178	84	-0.13	037.43	0
037	Zinc (ppm)	0009	185.0	6.000	187.3	30.14	6.178	84	-0.08	037.42	0
037	Zinc (ppm)	0870	185.3	1.600	187.3	30.14	6.178	84	-0.07	037.42	0
037	Zinc (ppm)	0242	187.0	12.00	187.3	30.14	6.178	84	-0.01	037.99	0
037	Zinc (ppm)	0208	187.8	4.100	187.3	30.14	6.178	84	0.01	037.41	0
037	Zinc (ppm)	0345	189.5	1.000	187.3	30.14	6.178	84	0.07	037.43	0
037	Zinc (ppm)	0682	191.0	0.0000	187.3	30.14	6.178	84	0.12	037.43	0
037	Zinc (ppm)	0939	191.5	3.000	187.3	30.14	6.178	84	0.14	037.53	0
037	Zinc (ppm)	0263	191.6	0.6770	187.3	30.14	6.178	84	0.14	037.42	0
037	Zinc (ppm)	0407	192.5	1.191	187.3	30.14	6.178	84	0.17	037.41	0
037	Zinc (ppm)	2129	193.0	4.000	187.3	30.14	6.178	84	0.19	037.42	0
037	Zinc (ppm)	2058	193.2	10.08	187.3	30.14	6.178	84	0.19	037.43	0
037	Zinc (ppm)	0278	194.1	4.900	187.3	30.14	6.178	84	0.22	037.42	0
037	Zinc (ppm)	0017	195.0	9.295	187.3	30.14	6.178	84	0.25	037.43	0
037	Zinc (ppm)	0357	196.1	1.684	187.3	30.14	6.178	84	0.29	037.42	0
037	Zinc (ppm)	0563	196.8	1.024	187.3	30.14	6.178	84	0.31	037.31	0
037	Zinc (ppm)	0033	197.0	4.000	187.3	30.14	6.178	84	0.32	037.43	0
037	Zinc (ppm)	0148	198.4	0.4300	187.3	30.14	6.178	84	0.37	037.41	0
037	Zinc (ppm)	0038	198.6	10.90	187.3	30.14	6.178	84	0.37	037.32	0
037	Zinc (ppm)	2319	199.0	6.000	187.3	30.14	6.178	84	0.39	037.42	0
037	Zinc (ppm)	0010	200.0	0.0000	187.3	30.14	6.178	84	0.42	037.33	0
037	Zinc (ppm)	0407	200.2	10.67	187.3	30.14	6.178	84	0.43	037.43	0
037	Zinc (ppm)	2342	200.5	7.000	187.3	30.14	6.178	84	0.44	037.44	0
037	Zinc (ppm)	2089	201.1	1.440	187.3	30.14	6.178	84	0.46	037.43	0
037	Zinc (ppm)	0265	201.5	5.000	187.3	30.14	6.178	84	0.47	037.42	0
037	Zinc (ppm)	0160	201.5	7.000	187.3	30.14	6.178	84	0.47	037.44	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
037	Zinc (ppm)	0083	202.5	7.000	187.3	30.14	6.178	84	0.50	037.43	0
037	Zinc (ppm)	0035	204.0	4.000	187.3	30.14	6.178	84	0.55	037.43	0
037	Zinc (ppm)	0510	204.0	10.00	187.3	30.14	6.178	84	0.55	037.43	0
037	Zinc (ppm)	0202	205.5	7.000	187.3	30.14	6.178	84	0.60	037.43	0
037	Zinc (ppm)	0353	206.5	5.000	187.3	30.14	6.178	84	0.64	037.43	0
037	Zinc (ppm)	0870	206.8	2.300	187.3	30.14	6.178	84	0.64	037.43	0
037	Zinc (ppm)	0208	207.5	11.00	187.3	30.14	6.178	84	0.67	037.31	0
037	Zinc (ppm)	0726	208.6	4.000	187.3	30.14	6.178	84	0.71	037.42	0
037	Zinc (ppm)	0675	208.8	5.640	187.3	30.14	6.178	84	0.71	037.43	0
037	Zinc (ppm)	0505	209.0	0.1000	187.3	30.14	6.178	84	0.72	037.43	0
037	Zinc (ppm)	0692	209.6	9.800	187.3	30.14	6.178	84	0.74	037.42	0
037	Zinc (ppm)	0100	210.5	13.00	187.3	30.14	6.178	84	0.77	037.99	0
037	Zinc (ppm)	2192	212.5	21.00	187.3	30.14	6.178	84	0.83	037.43	0
037	Zinc (ppm)	0169	214.0	14.00	187.3	30.14	6.178	84	0.88	037.43	0
037	Zinc (ppm)	0553	215.0	22.00	187.3	30.14	6.178	84	0.92	037.53	0
037	Zinc (ppm)	0425	215.7	5.700	187.3	30.14	6.178	84	0.94	037.43	0
037	Zinc (ppm)	0693	216.3	20.37	187.3	30.14	6.178	84	0.96	037.42	0
037	Zinc (ppm)	0027	217.3	3.980	187.3	30.14	6.178	84	0.99	037.43	0
037	Zinc (ppm)	0098	217.4	1.000	187.3	30.14	6.178	84	1.00	037.43	0
037	Zinc (ppm)	0042	217.5	3.000	187.3	30.14	6.178	84	1.00	037.42	0
037	Zinc (ppm)	0035	218.0	0.0000	187.3	30.14	6.178	84	1.02	037.42	0
037	Zinc (ppm)	2326	219.4	1.100	187.3	30.14	6.178	84	1.06	037.43	0
037	Zinc (ppm)	0190	225.3	4.000	187.3	30.14	6.178	84	1.26	037.42	0
037	Zinc (ppm)	0572	228.0	2.000	187.3	30.14	6.178	84	1.35	037.53	0
037	Zinc (ppm)	0037	236.9	20.10	187.3	30.14	6.178	84	1.64	037.42	0
037	Zinc (ppm)	0038	240.7	38.30	187.3	30.14	6.178	84	1.77	037.43	0
037	Zinc (ppm)	0407	245.4	7.161	187.3	30.14	6.178	84	1.93	037.53	0
037	Zinc (ppm)	0610	121.5	191.0	187.3	30.14	6.178	84	-2.18	037.43	1
037	Zinc (ppm)	0297	1,136	3.000	187.3	30.14	6.178	84	31.46	037.43	2
038	Molybdenum (ppm)	0345	1.950	0.1000	3.025	0.8559	0.0917	17	-1.26	038.43	0
038	Molybdenum (ppm)	0171	2.030	0.0880	3.025	0.8559	0.0917	17	-1.16	038.41	0
038	Molybdenum (ppm)	0083	2.100	0.0000	3.025	0.8559	0.0917	17	-1.08	038.43	0
038	Molybdenum (ppm)	0038	2.334	0.0310	3.025	0.8559	0.0917	17	-0.81	038.43	0
038	Molybdenum (ppm)	0407	2.425	0.2318	3.025	0.8559	0.0917	17	-0.70	038.43	0
038	Molybdenum (ppm)	0202	2.575	0.0700	3.025	0.8559	0.0917	17	-0.53	038.43	0
038	Molybdenum (ppm)	0226	2.750	0.0600	3.025	0.8559	0.0917	17	-0.32	038.41	0
038	Molybdenum (ppm)	0278	2.760	0.0000	3.025	0.8559	0.0917	17	-0.31	038.42	0
038	Molybdenum (ppm)	0407	2.849	0.0331	3.025	0.8559	0.0917	17	-0.21	038.41	0
038	Molybdenum (ppm)	0693	3.238	0.0580	3.025	0.8559	0.0917	17	0.25	038.42	0
038	Molybdenum (ppm)	0939	3.365	0.0500	3.025	0.8559	0.0917	17	0.40	038.53	0
038	Molybdenum (ppm)	0510	3.450	0.1000	3.025	0.8559	0.0917	17	0.50	038.43	0
038	Molybdenum (ppm)	0265	3.550	0.2800	3.025	0.8559	0.0917	17	0.61	038.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
038	Molybdenum (ppm)	0553	3.620	0.1200	3.025	0.8559	0.0917	17	0.69	038.53	0
038	Molybdenum (ppm)	0407	3.816	0.1869	3.025	0.8559	0.0917	17	0.92	038.53	0
038	Molybdenum (ppm)	0169	4.765	0.0700	3.025	0.8559	0.0917	17	2.03	038.43	0
038	Molybdenum (ppm)	0297	12.65	0.0800	3.025	0.8559	0.0917	17	11.25	038.43	0
038	Molybdenum (ppm)	0045	3.610	0.4800	3.025	0.8559	0.0917	17	0.68	038.42	1
040	Barium (ppm)	0407	23.49	2.198				1		040.53	0
041	Vanadium (ppm)	0553	1.920	0.0000				2		041.53	0
041	Vanadium (ppm)	0407	2.338	0.1446				2		041.53	0
042	Chloride (%)	2188	0.5550	0.0000	0.6363	0.0548	0.0050	4		042.01	0
042	Chloride (%)	0226	0.6550	0.0100	0.6363	0.0548	0.0050	4		042.00	0
042	Chloride (%)	0948	0.6600	0.0000	0.6363	0.0548	0.0050	4		042.00	0
042	Chloride (%)	2326	0.6750	0.0100	0.6363	0.0548	0.0050	4		042.00	0
042	Chloride (%)	0297	2.965	0.0900	0.6363	0.0548	0.0050	4		042.99	2
099	Menadione (form) (ppm)	2192	0.6000	0.1200				1		099.01	0
101	Choline Chloride (ppm)	0227	1,800	0.0000				1		101.99	0
102	Niacin (ppm)	2319	98.15	3.700				3		102.02	0
102	Niacin (ppm)	2325	120.8	1.480				3		102.02	0
102	Niacin (ppm)	0227	189.0	6.000				3		102.01	0
103	Pantothenic Acid (ppm)	0227	73.05	0.5000				2		103.01	0
103	Pantothenic Acid (ppm)	0010	80.80	8.200				2		103.02	0
104	Riboflavin (ppm)	0910	11.18	0.1800				2		104.03	0
104	Riboflavin (ppm)	0227	17.40	0.8000				2		104.00	0
105	Thiamine (ppm)	2319	14.70	0.0000				3		105.00	0
105	Thiamine (ppm)	0910	19.54	0.2600				3		105.00	0
105	Thiamine (ppm)	0227	28.05	0.1000				3		105.01	0
105	Thiamine (ppm)	0010	8.500	1.800				3		105.00	1
106	Vitamin A (KU / kg)	0038	4.971	1.552	9.262	1.953	1.356	11	-2.20	106.02	0
106	Vitamin A (KU / kg)	0910	6.500	1.800	9.262	1.953	1.356	11	-1.41	106.02	0
106	Vitamin A (KU / kg)	0148	8.550	0.3660	9.262	1.953	1.356	11	-0.36	106.02	0
106	Vitamin A (KU / kg)	0563	8.703	3.107	9.262	1.953	1.356	11	-0.29	106.02	0
106	Vitamin A (KU / kg)	0227	8.755	0.4500	9.262	1.953	1.356	11	-0.26	106.00	0
106	Vitamin A (KU / kg)	0010	9.675	2.290	9.262	1.953	1.356	11	0.21	106.02	0
106	Vitamin A (KU / kg)	0036	10.15	1.100	9.262	1.953	1.356	11	0.45	106.02	0
106	Vitamin A (KU / kg)	0004	10.23	0.0000	9.262	1.953	1.356	11	0.50	106.02	0
106	Vitamin A (KU / kg)	0610	10.25	3.500	9.262	1.953	1.356	11	0.51	106.02	0
106	Vitamin A (KU / kg)	2103	10.80	0.1100	9.262	1.953	1.356	11	0.78	106.02	0
106	Vitamin A (KU / kg)	0208	11.94	0.6390	9.262	1.953	1.356	11	1.37	106.02	0
106	Vitamin A (KU / kg)	0098	16.90	9.000	9.262	1.953	1.356	11	3.91	106.01	1
107	Vitamin B12 (ppb)	0227	18.35	0.5000				2		107.00	0
107	Vitamin B12 (ppb)	0160	170.0	60.00				2		107.99	0
108	Vitamin D3 (KU / kg)	2319	1.930	0.0600	2.178	0.3131	0.0350	4		108.02	0
108	Vitamin D3 (KU / kg)	0051	1.985	0.0300	2.178	0.3131	0.0350	4		108.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
108	Vitamin D3 (KU / kg)	0010	2.175	0.0100	2.178	0.3131	0.0350	4		108.99	0
108	Vitamin D3 (KU / kg)	0914	2.620	0.0400	2.178	0.3131	0.0350	4		108.02	0
108	Vitamin D3 (KU / kg)	0227	3.680	0.4000	2.178	0.3131	0.0350	4		108.02	1
109	Vitamin E (IU / kg)	0098	74.05	6.900	96.73	17.59	5.131	9	-1.29	109.02	0
109	Vitamin E (IU / kg)	0208	81.50	2.642	96.73	17.59	5.131	9	-0.87	109.02	0
109	Vitamin E (IU / kg)	0036	84.45	1.900	96.73	17.59	5.131	9	-0.70	109.02	0
109	Vitamin E (IU / kg)	0610	91.95	1.900	96.73	17.59	5.131	9	-0.27	109.02	0
109	Vitamin E (IU / kg)	0227	95.35	0.5000	96.73	17.59	5.131	9	-0.08	109.02	0
109	Vitamin E (IU / kg)	0148	105.2	22.20	96.73	17.59	5.131	9	0.48	109.02	0
109	Vitamin E (IU / kg)	0171	106.0	2.000	96.73	17.59	5.131	9	0.53	109.99	0
109	Vitamin E (IU / kg)	0563	108.9	6.719	96.73	17.59	5.131	9	0.69	109.02	0
109	Vitamin E (IU / kg)	2193	149.6	1.420	96.73	17.59	5.131	9	3.00	109.02	0
109	Vitamin E (IU / kg)	0038	251.0	62.00	96.73	17.59	5.131	9	8.77	109.02	1
111	Vitamin C, Phosphorylated (ppm)	0227	< 4.4					0		111.00	5
112	Pyridoxine (µg / g)	0010	17.00	0.6000				3		112.01	0
112	Pyridoxine (µg / g)	2319	17.60	0.0000				3		112.01	0
112	Pyridoxine (µg / g)	0227	19.20	0.2000				3		112.01	0
113	Folic Acid (ppm)	0227	4.715	0.0900				2		113.01	0
113	Folic Acid (ppm)	2319	4.800	0.0000				2		113.02	0
114	Biotin (ppm)	0227	0.6520	0.0100				1		114.01	0
115	Non Protein N (NPN) (%)	0098	0.6150	0.0100				1		115.00	0
118	Peroxide value (meq/kg)	2081	3.535	0.3700				1		118.99	0
120	Alanine (%)	0859	0.6684	0.0180	0.7828	0.0345	0.0096	23	-3.31	120.00	0
120	Alanine (%)	0407	0.7330	0.0100	0.7828	0.0345	0.0096	23	-1.44	120.05	0
120	Alanine (%)	0968	0.7400	0.0000	0.7828	0.0345	0.0096	23	-1.24	120.00	0
120	Alanine (%)	0353	0.7450	0.0100	0.7828	0.0345	0.0096	23	-1.10	120.00	0
120	Alanine (%)	0626	0.7600	0.0180	0.7828	0.0345	0.0096	23	-0.66	120.05	0
120	Alanine (%)	2319	0.7600	0.0000	0.7828	0.0345	0.0096	23	-0.66	120.05	0
120	Alanine (%)	2193	0.7640	0.0020	0.7828	0.0345	0.0096	23	-0.55	120.05	0
120	Alanine (%)	0354	0.7645	0.0010	0.7828	0.0345	0.0096	23	-0.53	120.00	0
120	Alanine (%)	2188	0.7700	0.0000	0.7828	0.0345	0.0096	23	-0.37	120.05	0
120	Alanine (%)	0682	0.7730	0.0000	0.7828	0.0345	0.0096	23	-0.28	120.00	0
120	Alanine (%)	0504	0.7800	0.0200	0.7828	0.0345	0.0096	23	-0.08	120.00	0
120	Alanine (%)	0652	0.7800	0.0000	0.7828	0.0345	0.0096	23	-0.08	120.00	0
120	Alanine (%)	0226	0.7927	0.0161	0.7828	0.0345	0.0096	23	0.28	120.00	0
120	Alanine (%)	0571	0.7935	0.0110	0.7828	0.0345	0.0096	23	0.31	120.00	0
120	Alanine (%)	2059	0.7935	0.0010	0.7828	0.0345	0.0096	23	0.31	120.00	0
120	Alanine (%)	0098	0.7950	0.0020	0.7828	0.0345	0.0096	23	0.35	120.02	0
120	Alanine (%)	2196	0.8080	0.0000	0.7828	0.0345	0.0096	23	0.73	120.05	0
120	Alanine (%)	0227	0.8100	0.0200	0.7828	0.0345	0.0096	23	0.79	120.99	0
120	Alanine (%)	0644	0.8135	0.0070	0.7828	0.0345	0.0096	23	0.89	120.00	0
120	Alanine (%)	0675	0.8150	0.0100	0.7828	0.0345	0.0096	23	0.93	120.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
120	Alanine (%)	0941	0.8240	0.0000	0.7828	0.0345	0.0096	23	1.19	120.00	0
120	Alanine (%)	0939	0.8250	0.0500	0.7828	0.0345	0.0096	23	1.22	120.00	0
120	Alanine (%)	0148	0.8425	0.0250	0.7828	0.0345	0.0096	23	1.73	120.05	0
120	Alanine (%)	0870	0.8009	0.0729	0.7828	0.0345	0.0096	23	0.52	120.00	1
121	Arginine (%)	0859	0.8735	0.0260	0.9713	0.0495	0.0164	23	-1.97	121.00	0
121	Arginine (%)	0407	0.8910	0.0000	0.9713	0.0495	0.0164	23	-1.62	121.05	0
121	Arginine (%)	0682	0.9120	0.0000	0.9713	0.0495	0.0164	23	-1.20	121.00	0
121	Arginine (%)	0504	0.9250	0.0100	0.9713	0.0495	0.0164	23	-0.93	121.00	0
121	Arginine (%)	0968	0.9300	0.0200	0.9713	0.0495	0.0164	23	-0.83	121.00	0
121	Arginine (%)	0354	0.9390	0.0040	0.9713	0.0495	0.0164	23	-0.65	121.00	0
121	Arginine (%)	2188	0.9400	0.0000	0.9713	0.0495	0.0164	23	-0.63	121.05	0
121	Arginine (%)	0098	0.9480	0.0040	0.9713	0.0495	0.0164	23	-0.47	121.02	0
121	Arginine (%)	0870	0.9541	0.1029	0.9713	0.0495	0.0164	23	-0.35	121.00	0
121	Arginine (%)	0227	0.9550	0.0100	0.9713	0.0495	0.0164	23	-0.33	121.99	0
121	Arginine (%)	2193	0.9710	0.0020	0.9713	0.0495	0.0164	23	-0.01	121.05	0
121	Arginine (%)	2319	0.9750	0.0100	0.9713	0.0495	0.0164	23	0.08	121.05	0
121	Arginine (%)	0353	0.9850	0.0300	0.9713	0.0495	0.0164	23	0.28	121.00	0
121	Arginine (%)	0652	0.9900	0.0000	0.9713	0.0495	0.0164	23	0.38	121.00	0
121	Arginine (%)	0675	0.9900	0.0200	0.9713	0.0495	0.0164	23	0.38	121.00	0
121	Arginine (%)	0939	0.9900	0.0200	0.9713	0.0495	0.0164	23	0.38	121.00	0
121	Arginine (%)	2059	1.000	0.0140	0.9713	0.0495	0.0164	23	0.58	121.00	0
121	Arginine (%)	0226	1.001	0.0038	0.9713	0.0495	0.0164	23	0.60	121.00	0
121	Arginine (%)	0644	1.007	0.0230	0.9713	0.0495	0.0164	23	0.71	121.00	0
121	Arginine (%)	0571	1.011	0.0170	0.9713	0.0495	0.0164	23	0.79	121.00	0
121	Arginine (%)	2196	1.032	0.0000	0.9713	0.0495	0.0164	23	1.23	121.05	0
121	Arginine (%)	0941	1.047	0.0000	0.9713	0.0495	0.0164	23	1.53	121.00	0
121	Arginine (%)	0148	1.090	0.0600	0.9713	0.0495	0.0164	23	2.40	121.05	0
121	Arginine (%)	0626	495.0	988.0	0.9713	0.0495	0.0164	23	9971.85	121.05	2
122	Aspartic (%)	0859	1.525	0.0392	1.738	0.0473	0.0242	23	-4.50	122.00	0
122	Aspartic (%)	0407	1.534	0.0540	1.738	0.0473	0.0242	23	-4.32	122.05	0
122	Aspartic (%)	0682	1.629	0.0000	1.738	0.0473	0.0242	23	-2.31	122.00	0
122	Aspartic (%)	0968	1.660	0.0000	1.738	0.0473	0.0242	23	-1.65	122.00	0
122	Aspartic (%)	0504	1.700	0.0400	1.738	0.0473	0.0242	23	-0.81	122.00	0
122	Aspartic (%)	2188	1.720	0.0000	1.738	0.0473	0.0242	23	-0.38	122.05	0
122	Aspartic (%)	0354	1.723	0.0080	1.738	0.0473	0.0242	23	-0.32	122.00	0
122	Aspartic (%)	2193	1.727	0.0020	1.738	0.0473	0.0242	23	-0.23	122.05	0
122	Aspartic (%)	0571	1.738	0.0340	1.738	0.0473	0.0242	23	0.00	122.00	0
122	Aspartic (%)	0226	1.738	0.0295	1.738	0.0473	0.0242	23	0.00	122.00	0
122	Aspartic (%)	0227	1.745	0.0100	1.738	0.0473	0.0242	23	0.15	122.99	0
122	Aspartic (%)	2059	1.746	0.0160	1.738	0.0473	0.0242	23	0.17	122.00	0
122	Aspartic (%)	2196	1.746	0.0000	1.738	0.0473	0.0242	23	0.17	122.05	0
122	Aspartic (%)	0353	1.750	0.0200	1.738	0.0473	0.0242	23	0.25	122.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
122	Aspartic (%)	2319	1.755	0.0100	1.738	0.0473	0.0242	23	0.36	122.05	0
122	Aspartic (%)	0644	1.756	0.0350	1.738	0.0473	0.0242	23	0.37	122.00	0
122	Aspartic (%)	0098	1.760	0.0330	1.738	0.0473	0.0242	23	0.45	122.02	0
122	Aspartic (%)	0941	1.766	0.0000	1.738	0.0473	0.0242	23	0.59	122.00	0
122	Aspartic (%)	0652	1.770	0.0200	1.738	0.0473	0.0242	23	0.68	122.00	0
122	Aspartic (%)	0939	1.770	0.0200	1.738	0.0473	0.0242	23	0.68	122.00	0
122	Aspartic (%)	0675	1.780	0.0400	1.738	0.0473	0.0242	23	0.89	122.00	0
122	Aspartic (%)	0148	1.990	0.0800	1.738	0.0473	0.0242	23	5.33	122.05	0
122	Aspartic (%)	0626	2.035	0.0670	1.738	0.0473	0.0242	23	6.27	122.05	0
122	Aspartic (%)	0870	1.790	0.1618	1.738	0.0473	0.0242	23	1.10	122.00	1
124	Cysteine/Cystine (%)	0407	0.1171	0.0079	0.2326	0.0336	0.0065	21	-3.44	124.05	0
124	Cysteine/Cystine (%)	0148	0.1580	0.0500	0.2326	0.0336	0.0065	21	-2.22	124.05	0
124	Cysteine/Cystine (%)	0870	0.1686	0.0105	0.2326	0.0336	0.0065	21	-1.91	124.00	0
124	Cysteine/Cystine (%)	0227	0.2000	0.0000	0.2326	0.0336	0.0065	21	-0.97	124.99	0
124	Cysteine/Cystine (%)	0859	0.2118	0.0039	0.2326	0.0336	0.0065	21	-0.62	124.00	0
124	Cysteine/Cystine (%)	0652	0.2150	0.0100	0.2326	0.0336	0.0065	21	-0.52	124.00	0
124	Cysteine/Cystine (%)	0675	0.2200	0.0200	0.2326	0.0336	0.0065	21	-0.38	124.00	0
124	Cysteine/Cystine (%)	2059	0.2200	0.0040	0.2326	0.0336	0.0065	21	-0.38	124.00	0
124	Cysteine/Cystine (%)	2319	0.2300	0.0000	0.2326	0.0336	0.0065	21	-0.08	124.05	0
124	Cysteine/Cystine (%)	0644	0.2305	0.0010	0.2326	0.0336	0.0065	21	-0.06	124.00	0
124	Cysteine/Cystine (%)	2196	0.2380	0.0000	0.2326	0.0336	0.0065	21	0.16	124.05	0
124	Cysteine/Cystine (%)	0571	0.2390	0.0040	0.2326	0.0336	0.0065	21	0.19	124.00	0
124	Cysteine/Cystine (%)	0939	0.2450	0.0100	0.2326	0.0336	0.0065	21	0.37	124.00	0
124	Cysteine/Cystine (%)	0941	0.2480	0.0000	0.2326	0.0336	0.0065	21	0.46	124.00	0
124	Cysteine/Cystine (%)	0226	0.2485	0.0008	0.2326	0.0336	0.0065	21	0.47	124.00	0
124	Cysteine/Cystine (%)	0968	0.2500	0.0000	0.2326	0.0336	0.0065	21	0.52	124.00	0
124	Cysteine/Cystine (%)	0098	0.2575	0.0030	0.2326	0.0336	0.0065	21	0.74	124.02	0
124	Cysteine/Cystine (%)	0504	0.2650	0.0100	0.2326	0.0336	0.0065	21	0.96	124.00	0
124	Cysteine/Cystine (%)	0354	0.2670	0.0020	0.2326	0.0336	0.0065	21	1.02	124.00	0
124	Cysteine/Cystine (%)	2188	0.2700	0.0000	0.2326	0.0336	0.0065	21	1.11	124.05	0
124	Cysteine/Cystine (%)	0682	0.2920	0.0000	0.2326	0.0336	0.0065	21	1.77	124.00	0
124	Cysteine/Cystine (%)	0353	0.2500	0.0600	0.2326	0.0336	0.0065	21	0.52	124.00	1
125	Glutamic (%)	0407	2.145	0.0270	2.349	0.1023	0.0353	23	-2.00	125.05	0
125	Glutamic (%)	0682	2.210	0.0000	2.349	0.1023	0.0353	23	-1.36	125.00	0
125	Glutamic (%)	0859	2.213	0.0652	2.349	0.1023	0.0353	23	-1.33	125.00	0
125	Glutamic (%)	0504	2.260	0.0600	2.349	0.1023	0.0353	23	-0.87	125.00	0
125	Glutamic (%)	0968	2.260	0.0000	2.349	0.1023	0.0353	23	-0.87	125.00	0
125	Glutamic (%)	2193	2.289	0.0010	2.349	0.1023	0.0353	23	-0.59	125.99	0
125	Glutamic (%)	2188	2.290	0.0000	2.349	0.1023	0.0353	23	-0.57	125.05	0
125	Glutamic (%)	2319	2.310	0.0200	2.349	0.1023	0.0353	23	-0.38	125.05	0
125	Glutamic (%)	0354	2.322	0.0220	2.349	0.1023	0.0353	23	-0.26	125.00	0
125	Glutamic (%)	0098	2.331	0.0340	2.349	0.1023	0.0353	23	-0.17	125.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
125	Glutamic (%)	0652	2.335	0.0100	2.349	0.1023	0.0353	23	-0.13	125.00	0
125	Glutamic (%)	2059	2.344	0.0160	2.349	0.1023	0.0353	23	-0.05	125.00	0
125	Glutamic (%)	0227	2.345	0.0500	2.349	0.1023	0.0353	23	-0.04	125.99	0
125	Glutamic (%)	0571	2.362	0.0110	2.349	0.1023	0.0353	23	0.13	125.00	0
125	Glutamic (%)	2196	2.379	0.0000	2.349	0.1023	0.0353	23	0.30	125.05	0
125	Glutamic (%)	0644	2.384	0.0490	2.349	0.1023	0.0353	23	0.34	125.00	0
125	Glutamic (%)	0675	2.405	0.0300	2.349	0.1023	0.0353	23	0.55	125.00	0
125	Glutamic (%)	0870	2.406	0.1978	2.349	0.1023	0.0353	23	0.56	125.00	0
125	Glutamic (%)	0941	2.440	0.0000	2.349	0.1023	0.0353	23	0.89	125.00	0
125	Glutamic (%)	0353	2.465	0.0100	2.349	0.1023	0.0353	23	1.14	125.00	0
125	Glutamic (%)	0626	2.488	0.1270	2.349	0.1023	0.0353	23	1.36	125.05	0
125	Glutamic (%)	0226	2.488	0.0609	2.349	0.1023	0.0353	23	1.36	125.00	0
125	Glutamic (%)	0939	2.500	0.0200	2.349	0.1023	0.0353	23	1.48	125.00	0
125	Glutamic (%)	0148	2.665	0.2300	2.349	0.1023	0.0353	23	3.09	125.05	1
126	Glycine (%)	0859	0.6715	0.0185	0.7685	0.0362	0.0152	24	-2.68	126.00	0
126	Glycine (%)	2193	0.7130	0.0740	0.7685	0.0362	0.0152	24	-1.53	126.05	0
126	Glycine (%)	2319	0.7150	0.0100	0.7685	0.0362	0.0152	24	-1.48	126.05	0
126	Glycine (%)	0968	0.7400	0.0000	0.7685	0.0362	0.0152	24	-0.79	126.00	0
126	Glycine (%)	0682	0.7430	0.0000	0.7685	0.0362	0.0152	24	-0.70	126.00	0
126	Glycine (%)	0354	0.7480	0.0040	0.7685	0.0362	0.0152	24	-0.57	126.00	0
126	Glycine (%)	2188	0.7500	0.0000	0.7685	0.0362	0.0152	24	-0.51	126.05	0
126	Glycine (%)	0407	0.7530	0.0320	0.7685	0.0362	0.0152	24	-0.43	126.05	0
126	Glycine (%)	0675	0.7550	0.0100	0.7685	0.0362	0.0152	24	-0.37	126.00	0
126	Glycine (%)	0504	0.7600	0.0200	0.7685	0.0362	0.0152	24	-0.23	126.00	0
126	Glycine (%)	0652	0.7600	0.0000	0.7685	0.0362	0.0152	24	-0.23	126.00	0
126	Glycine (%)	2196	0.7640	0.0000	0.7685	0.0362	0.0152	24	-0.12	126.05	0
126	Glycine (%)	0870	0.7660	0.0670	0.7685	0.0362	0.0152	24	-0.07	126.00	0
126	Glycine (%)	2059	0.7725	0.0010	0.7685	0.0362	0.0152	24	0.11	126.00	0
126	Glycine (%)	0939	0.7750	0.0100	0.7685	0.0362	0.0152	24	0.18	126.00	0
126	Glycine (%)	0353	0.7800	0.0000	0.7685	0.0362	0.0152	24	0.32	126.00	0
126	Glycine (%)	0571	0.7880	0.0000	0.7685	0.0362	0.0152	24	0.54	126.00	0
126	Glycine (%)	0226	0.7900	0.0059	0.7685	0.0362	0.0152	24	0.59	126.00	0
126	Glycine (%)	0644	0.7945	0.0110	0.7685	0.0362	0.0152	24	0.72	126.00	0
126	Glycine (%)	0227	0.7950	0.0100	0.7685	0.0362	0.0152	24	0.73	126.99	0
126	Glycine (%)	0098	0.8025	0.0050	0.7685	0.0362	0.0152	24	0.94	126.02	0
126	Glycine (%)	0941	0.8180	0.0000	0.7685	0.0362	0.0152	24	1.37	126.00	0
126	Glycine (%)	0626	0.8300	0.0320	0.7685	0.0362	0.0152	24	1.70	126.05	0
126	Glycine (%)	0148	0.8375	0.0550	0.7685	0.0362	0.0152	24	1.91	126.05	0
127	Histidine (%)	0859	0.3317	0.0050	0.3719	0.0191	0.0077	23	-2.11	127.00	0
127	Histidine (%)	0354	0.3425	0.0030	0.3719	0.0191	0.0077	23	-1.54	127.00	0
127	Histidine (%)	0968	0.3500	0.0000	0.3719	0.0191	0.0077	23	-1.15	127.00	0
127	Histidine (%)	2188	0.3500	0.0000	0.3719	0.0191	0.0077	23	-1.15	127.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
127	Histidine (%)	0226	0.3582	0.0014	0.3719	0.0191	0.0077	23	-0.72	127.00	0
127	Histidine (%)	2196	0.3590	0.0000	0.3719	0.0191	0.0077	23	-0.68	127.05	0
127	Histidine (%)	0098	0.3640	0.0000	0.3719	0.0191	0.0077	23	-0.41	127.02	0
127	Histidine (%)	0504	0.3650	0.0100	0.3719	0.0191	0.0077	23	-0.36	127.00	0
127	Histidine (%)	2319	0.3650	0.0100	0.3719	0.0191	0.0077	23	-0.36	127.05	0
127	Histidine (%)	2059	0.3660	0.0020	0.3719	0.0191	0.0077	23	-0.31	127.00	0
127	Histidine (%)	0571	0.3710	0.0020	0.3719	0.0191	0.0077	23	-0.05	127.00	0
127	Histidine (%)	0407	0.3730	0.0080	0.3719	0.0191	0.0077	23	0.06	127.05	0
127	Histidine (%)	0675	0.3750	0.0100	0.3719	0.0191	0.0077	23	0.16	127.00	0
127	Histidine (%)	0939	0.3750	0.0100	0.3719	0.0191	0.0077	23	0.16	127.00	0
127	Histidine (%)	0870	0.3764	0.0201	0.3719	0.0191	0.0077	23	0.23	127.00	0
127	Histidine (%)	0644	0.3775	0.0090	0.3719	0.0191	0.0077	23	0.29	127.00	0
127	Histidine (%)	0353	0.3850	0.0100	0.3719	0.0191	0.0077	23	0.69	127.00	0
127	Histidine (%)	0227	0.3850	0.0100	0.3719	0.0191	0.0077	23	0.69	127.99	0
127	Histidine (%)	0148	0.3875	0.0550	0.3719	0.0191	0.0077	23	0.82	127.05	0
127	Histidine (%)	0652	0.3900	0.0000	0.3719	0.0191	0.0077	23	0.95	127.00	0
127	Histidine (%)	0626	0.3935	0.0110	0.3719	0.0191	0.0077	23	1.13	127.05	0
127	Histidine (%)	0682	0.4010	0.0000	0.3719	0.0191	0.0077	23	1.52	127.00	0
127	Histidine (%)	0941	0.4110	0.0000	0.3719	0.0191	0.0077	23	2.05	127.00	0
127	Histidine (%)	2193	0.3315	0.0750	0.3719	0.0191	0.0077	23	-2.12	127.05	1
128	Isoleucine (%)	0354	0.5095	0.0030	0.6445	0.0395	0.0082	23	-3.41	128.00	0
128	Isoleucine (%)	2319	0.5150	0.0100	0.6445	0.0395	0.0082	23	-3.27	128.05	0
128	Isoleucine (%)	0682	0.5870	0.0000	0.6445	0.0395	0.0082	23	-1.45	128.00	0
128	Isoleucine (%)	0968	0.6000	0.0000	0.6445	0.0395	0.0082	23	-1.13	128.00	0
128	Isoleucine (%)	0652	0.6150	0.0100	0.6445	0.0395	0.0082	23	-0.75	128.00	0
128	Isoleucine (%)	0859	0.6170	0.0146	0.6445	0.0395	0.0082	23	-0.70	128.00	0
128	Isoleucine (%)	0353	0.6250	0.0100	0.6445	0.0395	0.0082	23	-0.49	128.00	0
128	Isoleucine (%)	0407	0.6295	0.0090	0.6445	0.0395	0.0082	23	-0.38	128.05	0
128	Isoleucine (%)	0939	0.6300	0.0000	0.6445	0.0395	0.0082	23	-0.37	128.00	0
128	Isoleucine (%)	2196	0.6490	0.0000	0.6445	0.0395	0.0082	23	0.11	128.05	0
128	Isoleucine (%)	0626	0.6530	0.0200	0.6445	0.0395	0.0082	23	0.21	128.05	0
128	Isoleucine (%)	2193	0.6540	0.0040	0.6445	0.0395	0.0082	23	0.24	128.05	0
128	Isoleucine (%)	0675	0.6550	0.0100	0.6445	0.0395	0.0082	23	0.27	128.00	0
128	Isoleucine (%)	0227	0.6550	0.0100	0.6445	0.0395	0.0082	23	0.27	128.99	0
128	Isoleucine (%)	0098	0.6555	0.0030	0.6445	0.0395	0.0082	23	0.28	128.02	0
128	Isoleucine (%)	0571	0.6575	0.0150	0.6445	0.0395	0.0082	23	0.33	128.00	0
128	Isoleucine (%)	2059	0.6580	0.0020	0.6445	0.0395	0.0082	23	0.34	128.00	0
128	Isoleucine (%)	0644	0.6720	0.0120	0.6445	0.0395	0.0082	23	0.70	128.00	0
128	Isoleucine (%)	0941	0.6730	0.0000	0.6445	0.0395	0.0082	23	0.72	128.00	0
128	Isoleucine (%)	0504	0.6800	0.0200	0.6445	0.0395	0.0082	23	0.90	128.00	0
128	Isoleucine (%)	2188	0.6800	0.0000	0.6445	0.0395	0.0082	23	0.90	128.05	0
128	Isoleucine (%)	0226	0.7274	0.0068	0.6445	0.0395	0.0082	23	2.10	128.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
128	Isoleucine (%)	0148	0.7350	0.0300	0.6445	0.0395	0.0082	23	2.29	128.05	0
128	Isoleucine (%)	0870	0.7121	0.1571	0.6445	0.0395	0.0082	23	1.71	128.00	1
129	Leucine (%)	0682	1.012	0.0000	1.107	0.0640	0.0143	23	-1.49	129.00	0
129	Leucine (%)	0859	1.019	0.0246	1.107	0.0640	0.0143	23	-1.39	129.00	0
129	Leucine (%)	0407	1.030	0.0130	1.107	0.0640	0.0143	23	-1.22	129.05	0
129	Leucine (%)	2319	1.035	0.0100	1.107	0.0640	0.0143	23	-1.13	129.05	0
129	Leucine (%)	0354	1.046	0.0070	1.107	0.0640	0.0143	23	-0.97	129.00	0
129	Leucine (%)	2193	1.058	0.0040	1.107	0.0640	0.0143	23	-0.77	129.05	0
129	Leucine (%)	0968	1.060	0.0000	1.107	0.0640	0.0143	23	-0.74	129.00	0
129	Leucine (%)	2196	1.066	0.0000	1.107	0.0640	0.0143	23	-0.65	129.05	0
129	Leucine (%)	0652	1.075	0.0100	1.107	0.0640	0.0143	23	-0.51	129.00	0
129	Leucine (%)	2059	1.125	0.0090	1.107	0.0640	0.0143	23	0.27	129.00	0
129	Leucine (%)	0675	1.125	0.0100	1.107	0.0640	0.0143	23	0.28	129.00	0
129	Leucine (%)	0626	1.127	0.0250	1.107	0.0640	0.0143	23	0.30	129.05	0
129	Leucine (%)	0504	1.130	0.0200	1.107	0.0640	0.0143	23	0.35	129.00	0
129	Leucine (%)	2188	1.130	0.0000	1.107	0.0640	0.0143	23	0.35	129.05	0
129	Leucine (%)	0571	1.134	0.0350	1.107	0.0640	0.0143	23	0.41	129.00	0
129	Leucine (%)	0939	1.140	0.0200	1.107	0.0640	0.0143	23	0.51	129.00	0
129	Leucine (%)	0644	1.145	0.0220	1.107	0.0640	0.0143	23	0.59	129.00	0
129	Leucine (%)	0227	1.145	0.0300	1.107	0.0640	0.0143	23	0.59	129.99	0
129	Leucine (%)	0098	1.151	0.0060	1.107	0.0640	0.0143	23	0.68	129.02	0
129	Leucine (%)	0941	1.162	0.0000	1.107	0.0640	0.0143	23	0.85	129.00	0
129	Leucine (%)	0226	1.169	0.0040	1.107	0.0640	0.0143	23	0.96	129.00	0
129	Leucine (%)	0353	1.185	0.0100	1.107	0.0640	0.0143	23	1.21	129.00	0
129	Leucine (%)	0148	1.245	0.0700	1.107	0.0640	0.0143	23	2.15	129.05	0
129	Leucine (%)	0870	1.116	0.0889	1.107	0.0640	0.0143	23	0.13	129.00	1
130	L-Lysine (%)	0407	0.7645	0.0410	0.8984	0.0365	0.0109	23	-3.67	130.05	0
130	L-Lysine (%)	0682	0.8270	0.0000	0.8984	0.0365	0.0109	23	-1.96	130.00	0
130	L-Lysine (%)	0859	0.8390	0.0185	0.8984	0.0365	0.0109	23	-1.63	130.00	0
130	L-Lysine (%)	2193	0.8645	0.0010	0.8984	0.0365	0.0109	23	-0.93	130.05	0
130	L-Lysine (%)	0354	0.8700	0.0080	0.8984	0.0365	0.0109	23	-0.78	130.00	0
130	L-Lysine (%)	0652	0.8700	0.0000	0.8984	0.0365	0.0109	23	-0.78	130.00	0
130	L-Lysine (%)	2319	0.8850	0.0100	0.8984	0.0365	0.0109	23	-0.37	130.05	0
130	L-Lysine (%)	0353	0.8900	0.0200	0.8984	0.0365	0.0109	23	-0.23	130.00	0
130	L-Lysine (%)	0571	0.8900	0.0260	0.8984	0.0365	0.0109	23	-0.23	130.00	0
130	L-Lysine (%)	0226	0.8935	0.0005	0.8984	0.0365	0.0109	23	-0.14	130.00	0
130	L-Lysine (%)	0968	0.9000	0.0000	0.8984	0.0365	0.0109	23	0.04	130.00	0
130	L-Lysine (%)	0148	0.9050	0.0300	0.8984	0.0365	0.0109	23	0.18	130.05	0
130	L-Lysine (%)	0644	0.9055	0.0150	0.8984	0.0365	0.0109	23	0.19	130.00	0
130	L-Lysine (%)	0870	0.9070	0.0399	0.8984	0.0365	0.0109	23	0.23	130.00	0
130	L-Lysine (%)	2059	0.9095	0.0030	0.8984	0.0365	0.0109	23	0.30	130.00	0
130	L-Lysine (%)	0939	0.9100	0.0000	0.8984	0.0365	0.0109	23	0.32	130.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
130	L-Lysine (%)	2188	0.9100	0.0000	0.8984	0.0365	0.0109	23	0.32	130.05	0
130	L-Lysine (%)	2196	0.9200	0.0000	0.8984	0.0365	0.0109	23	0.59	130.05	0
130	L-Lysine (%)	0941	0.9230	0.0000	0.8984	0.0365	0.0109	23	0.67	130.00	0
130	L-Lysine (%)	0626	0.9280	0.0180	0.8984	0.0365	0.0109	23	0.81	130.05	0
130	L-Lysine (%)	0504	0.9450	0.0100	0.8984	0.0365	0.0109	23	1.28	130.00	0
130	L-Lysine (%)	0098	0.9550	0.0000	0.8984	0.0365	0.0109	23	1.55	130.02	0
130	L-Lysine (%)	0675	0.9750	0.0100	0.8984	0.0365	0.0109	23	2.10	130.00	0
130	L-Lysine (%)	0227	0.9750	0.1100	0.8984	0.0365	0.0109	23	2.10	130.99	1
130	L-Lysine (%)	0848	0.3200	0.0000	0.8984	0.0365	0.0109	23	-15.86	130.00	2
131	Methionine (%)	0626	0.1995	0.0130	0.3235	0.0243	0.0122	23	-5.10	131.05	0
131	Methionine (%)	0407	0.2429	0.0214	0.3235	0.0243	0.0122	23	-3.31	131.05	0
131	Methionine (%)	0859	0.2945	0.0056	0.3235	0.0243	0.0122	23	-1.19	131.00	0
131	Methionine (%)	0682	0.3010	0.0000	0.3235	0.0243	0.0122	23	-0.92	131.00	0
131	Methionine (%)	0354	0.3095	0.0030	0.3235	0.0243	0.0122	23	-0.57	131.00	0
131	Methionine (%)	2188	0.3100	0.0000	0.3235	0.0243	0.0122	23	-0.55	131.05	0
131	Methionine (%)	2319	0.3100	0.0000	0.3235	0.0243	0.0122	23	-0.55	131.05	0
131	Methionine (%)	0939	0.3150	0.0300	0.3235	0.0243	0.0122	23	-0.35	131.00	0
131	Methionine (%)	0098	0.3160	0.0040	0.3235	0.0243	0.0122	23	-0.31	131.02	0
131	Methionine (%)	2196	0.3160	0.0000	0.3235	0.0243	0.0122	23	-0.31	131.05	0
131	Methionine (%)	0870	0.3185	0.0263	0.3235	0.0243	0.0122	23	-0.21	131.00	0
131	Methionine (%)	0941	0.3200	0.0000	0.3235	0.0243	0.0122	23	-0.14	131.00	0
131	Methionine (%)	0968	0.3200	0.0000	0.3235	0.0243	0.0122	23	-0.14	131.00	0
131	Methionine (%)	0148	0.3275	0.0550	0.3235	0.0243	0.0122	23	0.17	131.05	0
131	Methionine (%)	2059	0.3340	0.0040	0.3235	0.0243	0.0122	23	0.43	131.00	0
131	Methionine (%)	0504	0.3350	0.0100	0.3235	0.0243	0.0122	23	0.47	131.00	0
131	Methionine (%)	0652	0.3350	0.0100	0.3235	0.0243	0.0122	23	0.47	131.00	0
131	Methionine (%)	0571	0.3360	0.0040	0.3235	0.0243	0.0122	23	0.52	131.00	0
131	Methionine (%)	0226	0.3403	0.0044	0.3235	0.0243	0.0122	23	0.69	131.00	0
131	Methionine (%)	0644	0.3525	0.0010	0.3235	0.0243	0.0122	23	1.19	131.00	0
131	Methionine (%)	0675	0.3550	0.0100	0.3235	0.0243	0.0122	23	1.30	131.00	0
131	Methionine (%)	0227	0.3750	0.0300	0.3235	0.0243	0.0122	23	2.12	131.99	0
131	Methionine (%)	0353	0.4450	0.0500	0.3235	0.0243	0.0122	23	5.00	131.00	0
131	Methionine (%)	0848	0.9000	0.0000	0.3235	0.0243	0.0122	23	23.71	131.00	2
132	Phenylalanine (%)	0407	0.6305	0.0310	0.7324	0.0393	0.0172	24	-2.59	132.05	0
132	Phenylalanine (%)	0682	0.6490	0.0000	0.7324	0.0393	0.0172	24	-2.12	132.00	0
132	Phenylalanine (%)	0968	0.6900	0.0000	0.7324	0.0393	0.0172	24	-1.08	132.00	0
132	Phenylalanine (%)	0870	0.6969	0.0583	0.7324	0.0393	0.0172	24	-0.90	132.00	0
132	Phenylalanine (%)	0652	0.7050	0.0500	0.7324	0.0393	0.0172	24	-0.70	132.00	0
132	Phenylalanine (%)	2188	0.7100	0.0000	0.7324	0.0393	0.0172	24	-0.57	132.05	0
132	Phenylalanine (%)	2196	0.7140	0.0000	0.7324	0.0393	0.0172	24	-0.47	132.05	0
132	Phenylalanine (%)	2319	0.7150	0.0100	0.7324	0.0393	0.0172	24	-0.44	132.05	0
132	Phenylalanine (%)	0675	0.7200	0.0200	0.7324	0.0393	0.0172	24	-0.32	132.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
132	Phenylalanine (%)	0939	0.7200	0.0600	0.7324	0.0393	0.0172	24	-0.32	132.00	0
132	Phenylalanine (%)	0354	0.7210	0.0020	0.7324	0.0393	0.0172	24	-0.29	132.00	0
132	Phenylalanine (%)	2193	0.7235	0.0030	0.7324	0.0393	0.0172	24	-0.23	132.05	0
132	Phenylalanine (%)	0226	0.7242	0.0147	0.7324	0.0393	0.0172	24	-0.21	132.00	0
132	Phenylalanine (%)	2059	0.7350	0.0020	0.7324	0.0393	0.0172	24	0.07	132.00	0
132	Phenylalanine (%)	0098	0.7350	0.0020	0.7324	0.0393	0.0172	24	0.07	132.02	0
132	Phenylalanine (%)	0571	0.7415	0.0090	0.7324	0.0393	0.0172	24	0.23	132.00	0
132	Phenylalanine (%)	0626	0.7480	0.0140	0.7324	0.0393	0.0172	24	0.40	132.05	0
132	Phenylalanine (%)	0227	0.7550	0.0100	0.7324	0.0393	0.0172	24	0.58	132.99	0
132	Phenylalanine (%)	0644	0.7605	0.0170	0.7324	0.0393	0.0172	24	0.72	132.00	0
132	Phenylalanine (%)	0504	0.7650	0.0100	0.7324	0.0393	0.0172	24	0.83	132.00	0
132	Phenylalanine (%)	0859	0.7771	0.0157	0.7324	0.0393	0.0172	24	1.14	132.00	0
132	Phenylalanine (%)	0148	0.8275	0.0750	0.7324	0.0393	0.0172	24	2.42	132.05	0
132	Phenylalanine (%)	0941	0.8370	0.0000	0.7324	0.0393	0.0172	24	2.66	132.00	0
132	Phenylalanine (%)	0353	0.8450	0.0100	0.7324	0.0393	0.0172	24	2.87	132.00	0
133	Proline (%)	0354	0.8005	0.0050	0.8942	0.0537	0.0136	22	-1.74	133.00	0
133	Proline (%)	0870	0.8075	0.0523	0.8942	0.0537	0.0136	22	-1.62	133.00	0
133	Proline (%)	0407	0.8080	0.0060	0.8942	0.0537	0.0136	22	-1.61	133.05	0
133	Proline (%)	0571	0.8490	0.0040	0.8942	0.0537	0.0136	22	-0.84	133.00	0
133	Proline (%)	0968	0.8500	0.0000	0.8942	0.0537	0.0136	22	-0.82	133.00	0
133	Proline (%)	0652	0.8650	0.0100	0.8942	0.0537	0.0136	22	-0.54	133.00	0
133	Proline (%)	2319	0.8700	0.0000	0.8942	0.0537	0.0136	22	-0.45	133.05	0
133	Proline (%)	0682	0.8820	0.0000	0.8942	0.0537	0.0136	22	-0.23	133.00	0
133	Proline (%)	0939	0.8900	0.0200	0.8942	0.0537	0.0136	22	-0.08	133.00	0
133	Proline (%)	0644	0.8945	0.0190	0.8942	0.0537	0.0136	22	0.01	133.00	0
133	Proline (%)	0626	0.8985	0.0390	0.8942	0.0537	0.0136	22	0.08	133.05	0
133	Proline (%)	0859	0.9013	0.0202	0.8942	0.0537	0.0136	22	0.13	133.00	0
133	Proline (%)	0226	0.9037	0.0116	0.8942	0.0537	0.0136	22	0.18	133.00	0
133	Proline (%)	0504	0.9100	0.0200	0.8942	0.0537	0.0136	22	0.29	133.00	0
133	Proline (%)	2188	0.9100	0.0000	0.8942	0.0537	0.0136	22	0.29	133.05	0
133	Proline (%)	0227	0.9150	0.0100	0.8942	0.0537	0.0136	22	0.39	133.99	0
133	Proline (%)	2059	0.9155	0.0030	0.8942	0.0537	0.0136	22	0.40	133.00	0
133	Proline (%)	0941	0.9250	0.0000	0.8942	0.0537	0.0136	22	0.57	133.00	0
133	Proline (%)	0148	0.9500	0.0600	0.8942	0.0537	0.0136	22	1.04	133.05	0
133	Proline (%)	2196	0.9530	0.0000	0.8942	0.0537	0.0136	22	1.09	133.05	0
133	Proline (%)	0353	0.9850	0.0100	0.8942	0.0537	0.0136	22	1.69	133.00	0
133	Proline (%)	0675	0.9950	0.0100	0.8942	0.0537	0.0136	22	1.88	133.00	0
134	Serine (%)	0407	0.5320	0.0100	0.7494	0.0524	0.0081	22	-4.15	134.05	0
134	Serine (%)	0504	0.6350	0.0100	0.7494	0.0524	0.0081	22	-2.19	134.00	0
134	Serine (%)	0859	0.6802	0.0195	0.7494	0.0524	0.0081	22	-1.32	134.00	0
134	Serine (%)	0675	0.6850	0.0100	0.7494	0.0524	0.0081	22	-1.23	134.00	0
134	Serine (%)	0098	0.7090	0.0160	0.7494	0.0524	0.0081	22	-0.77	134.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
134	Serine (%)	0682	0.7220	0.0000	0.7494	0.0524	0.0081	22	-0.52	134.00	0
134	Serine (%)	2193	0.7305	0.0010	0.7494	0.0524	0.0081	22	-0.36	134.05	0
134	Serine (%)	0571	0.7365	0.0010	0.7494	0.0524	0.0081	22	-0.25	134.00	0
134	Serine (%)	0227	0.7450	0.0100	0.7494	0.0524	0.0081	22	-0.08	134.99	0
134	Serine (%)	0644	0.7460	0.0140	0.7494	0.0524	0.0081	22	-0.07	134.00	0
134	Serine (%)	0968	0.7500	0.0000	0.7494	0.0524	0.0081	22	0.01	134.00	0
134	Serine (%)	2188	0.7500	0.0000	0.7494	0.0524	0.0081	22	0.01	134.05	0
134	Serine (%)	0226	0.7567	0.0048	0.7494	0.0524	0.0081	22	0.14	134.00	0
134	Serine (%)	2059	0.7570	0.0060	0.7494	0.0524	0.0081	22	0.14	134.00	0
134	Serine (%)	0652	0.7650	0.0100	0.7494	0.0524	0.0081	22	0.30	134.00	0
134	Serine (%)	0941	0.7830	0.0000	0.7494	0.0524	0.0081	22	0.64	134.00	0
134	Serine (%)	0939	0.7900	0.0200	0.7494	0.0524	0.0081	22	0.78	134.00	0
134	Serine (%)	2196	0.7960	0.0000	0.7494	0.0524	0.0081	22	0.89	134.05	0
134	Serine (%)	0353	0.8050	0.0100	0.7494	0.0524	0.0081	22	1.06	134.00	0
134	Serine (%)	0354	0.8080	0.0060	0.7494	0.0524	0.0081	22	1.12	134.00	0
134	Serine (%)	2319	0.8100	0.0000	0.7494	0.0524	0.0081	22	1.16	134.05	0
134	Serine (%)	0626	0.8205	0.0290	0.7494	0.0524	0.0081	22	1.36	134.05	0
134	Serine (%)	0870	0.8353	0.0755	0.7494	0.0524	0.0081	22	1.64	134.00	1
134	Serine (%)	0148	0.8900	0.0800	0.7494	0.0524	0.0081	22	2.69	134.05	1
135	Threonine (%)	0407	0.5100	0.0020	0.6831	0.0253	0.0076	23	-6.84	135.05	0
135	Threonine (%)	0859	0.6069	0.0166	0.6831	0.0253	0.0076	23	-3.01	135.00	0
135	Threonine (%)	0675	0.6550	0.0100	0.6831	0.0253	0.0076	23	-1.11	135.00	0
135	Threonine (%)	0504	0.6600	0.0000	0.6831	0.0253	0.0076	23	-0.91	135.00	0
135	Threonine (%)	2188	0.6600	0.0000	0.6831	0.0253	0.0076	23	-0.91	135.05	0
135	Threonine (%)	0682	0.6630	0.0000	0.6831	0.0253	0.0076	23	-0.79	135.00	0
135	Threonine (%)	0226	0.6710	0.0042	0.6831	0.0253	0.0076	23	-0.48	135.00	0
135	Threonine (%)	2193	0.6740	0.0000	0.6831	0.0253	0.0076	23	-0.36	135.05	0
135	Threonine (%)	0098	0.6785	0.0070	0.6831	0.0253	0.0076	23	-0.18	135.02	0
135	Threonine (%)	0968	0.6800	0.0000	0.6831	0.0253	0.0076	23	-0.12	135.00	0
135	Threonine (%)	0644	0.6815	0.0130	0.6831	0.0253	0.0076	23	-0.06	135.00	0
135	Threonine (%)	2059	0.6825	0.0050	0.6831	0.0253	0.0076	23	-0.02	135.00	0
135	Threonine (%)	0939	0.6850	0.0300	0.6831	0.0253	0.0076	23	0.08	135.00	0
135	Threonine (%)	0571	0.6855	0.0190	0.6831	0.0253	0.0076	23	0.10	135.00	0
135	Threonine (%)	0652	0.6900	0.0200	0.6831	0.0253	0.0076	23	0.27	135.00	0
135	Threonine (%)	2319	0.6900	0.0000	0.6831	0.0253	0.0076	23	0.27	135.05	0
135	Threonine (%)	0626	0.6930	0.0320	0.6831	0.0253	0.0076	23	0.39	135.05	0
135	Threonine (%)	0227	0.7000	0.0000	0.6831	0.0253	0.0076	23	0.67	135.99	0
135	Threonine (%)	0941	0.7040	0.0000	0.6831	0.0253	0.0076	23	0.83	135.00	0
135	Threonine (%)	2196	0.7090	0.0000	0.6831	0.0253	0.0076	23	1.03	135.05	0
135	Threonine (%)	0354	0.7170	0.0000	0.6831	0.0253	0.0076	23	1.34	135.00	0
135	Threonine (%)	0353	0.7200	0.0000	0.6831	0.0253	0.0076	23	1.46	135.00	0
135	Threonine (%)	0148	0.7475	0.0150	0.6831	0.0253	0.0076	23	2.55	135.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
135	Threonine (%)	0870	0.7050	0.0697	0.6831	0.0253	0.0076	23	0.87	135.00	1
136	Tryptophan (%)	0098	0.1945	0.0010	0.2189	0.0190	0.0048	12	-1.28	136.02	0
136	Tryptophan (%)	0941	0.1970	0.0000	0.2189	0.0190	0.0048	12	-1.15	136.00	0
136	Tryptophan (%)	0407	0.2055	0.0030	0.2189	0.0190	0.0048	12	-0.70	136.05	0
136	Tryptophan (%)	2059	0.2120	0.0040	0.2189	0.0190	0.0048	12	-0.36	136.03	0
136	Tryptophan (%)	0859	0.2128	0.0140	0.2189	0.0190	0.0048	12	-0.32	136.03	0
136	Tryptophan (%)	0939	0.2150	0.0100	0.2189	0.0190	0.0048	12	-0.20	136.03	0
136	Tryptophan (%)	0682	0.2180	0.0000	0.2189	0.0190	0.0048	12	-0.05	136.00	0
136	Tryptophan (%)	0644	0.2200	0.0000	0.2189	0.0190	0.0048	12	0.06	136.01	0
136	Tryptophan (%)	0571	0.2290	0.0000	0.2189	0.0190	0.0048	12	0.54	136.03	0
136	Tryptophan (%)	0870	0.2302	0.0150	0.2189	0.0190	0.0048	12	0.60	136.00	0
136	Tryptophan (%)	0353	0.2450	0.0100	0.2189	0.0190	0.0048	12	1.38	136.03	0
136	Tryptophan (%)	0227	0.2700	0.0000	0.2189	0.0190	0.0048	12	2.70	136.00	0
136	Tryptophan (%)	0504	0.2100	0.0400	0.2189	0.0190	0.0048	12	-0.47	136.99	1
137	Tyrosine (%)	0354	0.3485	0.0030	0.4597	0.0765	0.0080	20	-1.45	137.00	0
137	Tyrosine (%)	0682	0.3530	0.0000	0.4597	0.0765	0.0080	20	-1.40	137.00	0
137	Tyrosine (%)	0675	0.3950	0.0100	0.4597	0.0765	0.0080	20	-0.85	137.00	0
137	Tyrosine (%)	0626	0.4040	0.0000	0.4597	0.0765	0.0080	20	-0.73	137.05	0
137	Tyrosine (%)	0870	0.4047	0.0229	0.4597	0.0765	0.0080	20	-0.72	137.00	0
137	Tyrosine (%)	2319	0.4050	0.0100	0.4597	0.0765	0.0080	20	-0.72	137.05	0
137	Tyrosine (%)	0968	0.4100	0.0000	0.4597	0.0765	0.0080	20	-0.65	137.00	0
137	Tyrosine (%)	0407	0.4260	0.0080	0.4597	0.0765	0.0080	20	-0.44	137.05	0
137	Tyrosine (%)	2196	0.4410	0.0000	0.4597	0.0765	0.0080	20	-0.24	137.05	0
137	Tyrosine (%)	2193	0.4495	0.0010	0.4597	0.0765	0.0080	20	-0.13	137.05	0
137	Tyrosine (%)	0227	0.4500	0.0000	0.4597	0.0765	0.0080	20	-0.13	137.99	0
137	Tyrosine (%)	0504	0.4850	0.0100	0.4597	0.0765	0.0080	20	0.33	137.00	0
137	Tyrosine (%)	0644	0.4955	0.0110	0.4597	0.0765	0.0080	20	0.47	137.00	0
137	Tyrosine (%)	0098	0.4970	0.0040	0.4597	0.0765	0.0080	20	0.49	137.02	0
137	Tyrosine (%)	0939	0.5000	0.0400	0.4597	0.0765	0.0080	20	0.53	137.00	0
137	Tyrosine (%)	0226	0.5065	0.0072	0.4597	0.0765	0.0080	20	0.61	137.00	0
137	Tyrosine (%)	2188	0.5300	0.0000	0.4597	0.0765	0.0080	20	0.92	137.05	0
137	Tyrosine (%)	0353	0.5550	0.0300	0.4597	0.0765	0.0080	20	1.25	137.00	0
137	Tyrosine (%)	0941	0.5640	0.0000	0.4597	0.0765	0.0080	20	1.36	137.00	0
137	Tyrosine (%)	2059	0.5920	0.0020	0.4597	0.0765	0.0080	20	1.73	137.00	0
137	Tyrosine (%)	0148	0.6650	0.0500	0.4597	0.0765	0.0080	20	2.68	137.05	1
138	Valine (%)	2319	0.6900	0.0000	0.8024	0.0519	0.0116	23	-2.16	138.05	0
138	Valine (%)	0354	0.6990	0.0040	0.8024	0.0519	0.0116	23	-1.99	138.00	0
138	Valine (%)	0870	0.7312	0.0570	0.8024	0.0519	0.0116	23	-1.37	138.00	0
138	Valine (%)	2196	0.7550	0.0000	0.8024	0.0519	0.0116	23	-0.91	138.05	0
138	Valine (%)	0941	0.7630	0.0000	0.8024	0.0519	0.0116	23	-0.76	138.00	0
138	Valine (%)	0407	0.7730	0.0080	0.8024	0.0519	0.0116	23	-0.57	138.05	0
138	Valine (%)	0652	0.7750	0.0100	0.8024	0.0519	0.0116	23	-0.53	138.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
138	Valine (%)	0682	0.7840	0.0000	0.8024	0.0519	0.0116	23	-0.35	138.00	0
138	Valine (%)	2193	0.7885	0.0050	0.8024	0.0519	0.0116	23	-0.27	138.05	0
138	Valine (%)	0968	0.7900	0.0000	0.8024	0.0519	0.0116	23	-0.24	138.00	0
138	Valine (%)	2059	0.8035	0.0070	0.8024	0.0519	0.0116	23	0.02	138.00	0
138	Valine (%)	0939	0.8050	0.0100	0.8024	0.0519	0.0116	23	0.05	138.00	0
138	Valine (%)	2188	0.8100	0.0000	0.8024	0.0519	0.0116	23	0.15	138.05	0
138	Valine (%)	0626	0.8130	0.0160	0.8024	0.0519	0.0116	23	0.20	138.05	0
138	Valine (%)	0504	0.8150	0.0300	0.8024	0.0519	0.0116	23	0.24	138.00	0
138	Valine (%)	0571	0.8160	0.0240	0.8024	0.0519	0.0116	23	0.26	138.00	0
138	Valine (%)	0098	0.8195	0.0110	0.8024	0.0519	0.0116	23	0.33	138.02	0
138	Valine (%)	0675	0.8400	0.0200	0.8024	0.0519	0.0116	23	0.72	138.00	0
138	Valine (%)	0644	0.8435	0.0110	0.8024	0.0519	0.0116	23	0.79	138.00	0
138	Valine (%)	0353	0.8500	0.0000	0.8024	0.0519	0.0116	23	0.92	138.00	0
138	Valine (%)	0227	0.8700	0.0200	0.8024	0.0519	0.0116	23	1.30	138.99	0
138	Valine (%)	0226	0.9034	0.0079	0.8024	0.0519	0.0116	23	1.94	138.00	0
138	Valine (%)	0148	0.9125	0.0250	0.8024	0.0519	0.0116	23	2.12	138.05	0
138	Valine (%)	0859	0.5673	0.3332	0.8024	0.0519	0.0116	23	-4.53	138.00	1
139	Taurine (%)	0407	0.0070	0.0000				3		139.05	0
139	Taurine (%)	0682	0.0630	0.0000				3		139.00	0
139	Taurine (%)	0504	0.1650	0.0300				3		139.00	0
139	Taurine (%)	0098	< 0.01					3		139.02	5
139	Taurine (%)	0227	< 0.01					3		139.99	5
160	Fructose (%)	0297	0.0400	0.0020				2		160.10	0
160	Fructose (%)	0227	0.8400	0.0000				2		160.99	0
161	Galactose (%)	0297	0.0000	0.0000				0		161.10	4
162	Glucose (%)	0297	0.0240	0.0120				2		162.10	0
162	Glucose (%)	0227	0.2900	0.0000				2		162.99	0
163	Lactose (%)	0297	0.0000	0.0000				0		163.10	4
163	Lactose (%)	0227	< 0.15					0		163.99	5
164	Maltose (%)	0297	0.1015	0.0090				2		164.10	0
164	Maltose (%)	0227	0.2350	0.0300				2		164.99	0
165	Sucrose (%)	0297	1.571	0.0080				2		165.10	0
165	Sucrose (%)	0227	2.235	0.0700				2		165.99	0
166	Raffinose (%)	0227	0.4000	0.0200				2		166.99	0
166	Raffinose (%)	0297	2.369	0.0300				2		166.10	0
167	Stachyose (%)	0227	0.6650	0.0300				2		167.99	0
167	Stachyose (%)	0297	477.5	953.0				2		167.10	0
351	Chlortetracycline (ppm)	0242	1.750	0.0600				3		351.05	0
351	Chlortetracycline (ppm)	0148	2.110	0.0600				3		351.05	0
351	Chlortetracycline (ppm)	0218	6.054	0.5610				3		351.00	0
386	Tiamulin (ppm)	0910	2.250	0.1000				1		386.99	0
388	Tylosin (ppm)	0875	1.800	0.2000				1		388.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
400	Water Activity (Units)	2268	0.3795	0.0050	0.5246	0.0153	0.0034	14	-9.46	400.01	0
400	Water Activity (Units)	2073	0.5070	0.0060	0.5246	0.0153	0.0034	14	-1.15	400.99	0
400	Water Activity (Units)	2303	0.5134	0.0121	0.5246	0.0153	0.0034	14	-0.73	400.01	0
400	Water Activity (Units)	2174	0.5145	0.0070	0.5246	0.0153	0.0034	14	-0.66	400.01	0
400	Water Activity (Units)	2168	0.5195	0.0070	0.5246	0.0153	0.0034	14	-0.33	400.01	0
400	Water Activity (Units)	0942	0.5210	0.0000	0.5246	0.0153	0.0034	14	-0.23	400.01	0
400	Water Activity (Units)	0407	0.5218	0.0020	0.5246	0.0153	0.0034	14	-0.18	400.01	0
400	Water Activity (Units)	0843	0.5250	0.0000	0.5246	0.0153	0.0034	14	0.03	400.99	0
400	Water Activity (Units)	2193	0.5260	0.0000	0.5246	0.0153	0.0034	14	0.09	400.01	0
400	Water Activity (Units)	0843	0.5300	0.0032	0.5246	0.0153	0.0034	14	0.35	400.01	0
400	Water Activity (Units)	0589	0.5320	0.0040	0.5246	0.0153	0.0034	14	0.48	400.01	0
400	Water Activity (Units)	0123	0.5400	0.0000	0.5246	0.0153	0.0034	14	1.00	400.01	0
400	Water Activity (Units)	2109	0.5450	0.0017	0.5246	0.0153	0.0034	14	1.33	400.01	0
400	Water Activity (Units)	0668	0.5500	0.0000	0.5246	0.0153	0.0034	14	1.66	400.01	0
516	Arsenic, Total (As) (ppm)	0045	0.1780	0.0000	0.2347	0.0466	0.0137	6	-1.22	516.00	0
516	Arsenic, Total (As) (ppm)	0171	0.2180	0.0140	0.2347	0.0466	0.0137	6	-0.36	516.00	0
516	Arsenic, Total (As) (ppm)	0227	0.2215	0.0250	0.2347	0.0466	0.0137	6	-0.28	516.53	0
516	Arsenic, Total (As) (ppm)	0553	0.2255	0.0050	0.2347	0.0466	0.0137	6	-0.20	516.53	0
516	Arsenic, Total (As) (ppm)	0939	0.2750	0.0300	0.2347	0.0466	0.0137	6	0.87	516.53	0
516	Arsenic, Total (As) (ppm)	0407	0.2900	0.0084	0.2347	0.0466	0.0137	6	1.19	516.53	0
516	Arsenic, Total (As) (ppm)	0682	< 10		0.2347	0.0466	0.0137	6		516.43	5
518	Cadmium (ppm)	0171	0.0810	0.0000	0.1199	0.0324	0.0124	7	-1.20	518.41	0
518	Cadmium (ppm)	0407	0.0834	0.0030	0.1199	0.0324	0.0124	7	-1.13	518.41	0
518	Cadmium (ppm)	0227	0.1130	0.0220	0.1199	0.0324	0.0124	7	-0.21	518.53	0
518	Cadmium (ppm)	0553	0.1325	0.0050	0.1199	0.0324	0.0124	7	0.39	518.53	0
518	Cadmium (ppm)	0939	0.1350	0.0300	0.1199	0.0324	0.0124	7	0.47	518.53	0
518	Cadmium (ppm)	0407	0.1400	0.0040	0.1199	0.0324	0.0124	7	0.62	518.43	0
518	Cadmium (ppm)	0407	0.1546	0.0229	0.1199	0.0324	0.0124	7	1.07	518.53	0
518	Cadmium (ppm)	0682	< 0.2		0.1199	0.0324	0.0124	7		518.43	5
520	Chromium, Total (Cr) (ppm)	0968	8.947	1.071	20.41	9.506	0.5074	11	-1.21	520.43	0
520	Chromium, Total (Cr) (ppm)	2325	10.11	0.2000	20.41	9.506	0.5074	11	-1.08	520.41	0
520	Chromium, Total (Cr) (ppm)	0171	10.89	0.4700	20.41	9.506	0.5074	11	-1.00	520.41	0
520	Chromium, Total (Cr) (ppm)	0407	14.39	1.050	20.41	9.506	0.5074	11	-0.63	520.43	0
520	Chromium, Total (Cr) (ppm)	0407	15.16	0.0530	20.41	9.506	0.5074	11	-0.55	520.41	0
520	Chromium, Total (Cr) (ppm)	0045	25.10	0.4000	20.41	9.506	0.5074	11	0.49	520.42	0
520	Chromium, Total (Cr) (ppm)	0510	27.35	0.5400	20.41	9.506	0.5074	11	0.73	520.43	0
520	Chromium, Total (Cr) (ppm)	0939	27.57	0.5200	20.41	9.506	0.5074	11	0.75	520.53	0
520	Chromium, Total (Cr) (ppm)	0407	27.62	0.5772	20.41	9.506	0.5074	11	0.76	520.53	0
520	Chromium, Total (Cr) (ppm)	0682	28.44	0.0000	20.41	9.506	0.5074	11	0.84	520.43	0
520	Chromium, Total (Cr) (ppm)	0553	28.95	0.7000	20.41	9.506	0.5074	11	0.90	520.53	0
520	Chromium, Total (Cr) (ppm)	0693	25.64	5.745	20.41	9.506	0.5074	11	0.55	520.42	1
526	Lead (ppm)	0171	0.1450	0.0000	0.2687	0.1107	0.0231	8	-1.12	526.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
526	Lead (ppm)	0407	0.1501	0.0112	0.2687	0.1107	0.0231	8	-1.07	526.41	0
526	Lead (ppm)	0227	0.2315	0.0290	0.2687	0.1107	0.0231	8	-0.34	526.53	0
526	Lead (ppm)	0553	0.2535	0.0110	0.2687	0.1107	0.0231	8	-0.14	526.53	0
526	Lead (ppm)	0939	0.2700	0.0600	0.2687	0.1107	0.0231	8	0.01	526.53	0
526	Lead (ppm)	0407	0.3270	0.0253	0.2687	0.1107	0.0231	8	0.53	526.53	0
526	Lead (ppm)	0407	0.3377	0.0483	0.2687	0.1107	0.0231	8	0.62	526.43	0
526	Lead (ppm)	0682	4.030	0.0000	0.2687	0.1107	0.0231	8	33.98	526.43	0
529	Mercury (ppb)	0968	1.690	0.2000				2		529.99	0
529	Mercury (ppb)	0407	4.809	0.5154				2		529.99	0
529	Mercury (ppb)	0227	< 10					2		529.99	5
529	Mercury (ppb)	0682	< 10000					2		529.99	5
539	Nickel (ppm)	0171	8.655	0.1900	12.76	2.759	0.2026	6	-1.49	539.41	0
539	Nickel (ppm)	0407	11.08	0.3180	12.76	2.759	0.2026	6	-0.61	539.41	0
539	Nickel (ppm)	0939	13.50	0.3100	12.76	2.759	0.2026	6	0.27	539.53	0
539	Nickel (ppm)	0407	13.83	0.1976	12.76	2.759	0.2026	6	0.39	539.53	0
539	Nickel (ppm)	0682	14.27	0.0000	12.76	2.759	0.2026	6	0.55	539.43	0
539	Nickel (ppm)	0553	15.20	0.2000	12.76	2.759	0.2026	6	0.89	539.53	0
539	Nickel (ppm)	0407	9.592	4.521	12.76	2.759	0.2026	6	-1.15	539.43	1
710	Lauric Acid (12:0) (% (w/w))	0226	< 0.005					0		710.99	5
710	Lauric Acid (12:0) (% (w/w))	0676	< 0.01					0		710.99	5
714	Myristic Acid (14:0) (% (w/w))	0226	0.0126	0.0001				1		714.99	0
716	Palmitic Acid (16:0) (% (w/w))	0226	0.6349	0.0018				1		716.99	0
718	Palmitoleic Acid (9c-16:1) (% (w/w))	0676	0.0065	0.0010				2		718.99	0
718	Palmitoleic Acid (9c-16:1) (% (w/w))	0226	0.0136	0.0005				2		718.99	0
722	Stearic Acid (18:0) (% (w/w))	0226	0.0980	0.0006				1		722.99	0
724	Oleic Acid (9c-18:1) (% (w/w))	0226	0.9454	0.0022				1		724.99	0
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0226	1.383	0.0058				2		726.99	0
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0676	1.528	0.0020				2		726.99	0
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/w))	0226	0.1444	0.0004				2		728.99	0
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/w))	0676	0.1555	0.0010				2		728.99	0
730	Arachidic Acid (20:0) (% (w/w))	0226	0.0249	0.0000				1		730.99	0
732	Gondoic Acid (11c-20:1) (% (w/w))	0226	0.0168	0.0014				1		732.99	0
736	Arachidonic Acid (5c,8c,11c,14c-20:4) (% (w/w))	0676	0.0000	0.0000				0		736.99	4
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c) (% (w/w))	0226	0.0000	0.0000				0		740.99	4
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c) (% (w/w))	0676	0.0000	0.0000				0		740.99	4
742	Behenic Acid (22:0) (% (w/w))	0226	0.0283	0.0001				1		742.99	0
744	Erucic Acid (13c-22:1) (% (w/w))	0226	< 0.005					0		744.99	5
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c) (% (w/w))	0676	0.0000	0.0000				0		746.99	4
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c) (% (w/w))	0226	< 0.005					0		746.99	5
748	Lignoceric Acid (24:0) (% (w/w))	0226	0.0329	0.0016				1		748.99	0
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22c) (% (w/w))	0226	0.0000	0.0000				0		750.99	4
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22c) (% (w/w))	0676	0.0000	0.0000				0		750.99	4

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
752	Nervonic Acid (24:1) isomers (% (w/w))	0226	< 0.005						0	752.99	5
754	Total n-3 Polyunsaturated (Omega-3) Fatty Aci	0676	0.1600	0.0000					1	754.99	0
756	Total n-6 Polyunsaturated (Omega-6) Fatty Aci	0676	1.530	0.0000					1	756.99	0
758	Total Saturated Fatty Acids (% (w/w))	0676	3.600	0.0000					1	758.99	0
772	Total Fatty Acids (% (w/w))	0226	3.335	0.0100					1	772.99	0

Notes: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2 and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = Rejected for duplicates too far apart, 2 = Rejected as outlier, 5 = A reporting limit and 4 = zeros submitted as values. Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). Flag 3 indicates not used in statistics.