



AAFCO
Proficiency Testing Program



Animal Feed Scheme

Soy Bean Meal

Test Material Code # 202129

Labs Reporting: 165

Analytes Reported 100

Issue Date : 10/31/2021

Analyte Proficiency Testing Report

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
000	Urea (%)	0278	0.2000	0.0000				1		000.02	0
001	Loss on Drying (%)	0618	8.955	0.0100	11.14	0.3859	0.1146	70	-5.67	001.99	0
001	Loss on Drying (%)	0019	9.650	0.3000	11.14	0.3859	0.1146	70	-3.87	001.07	0
001	Loss on Drying (%)	0612	9.945	0.7700	11.14	0.3859	0.1146	70	-3.10	001.99	0
001	Loss on Drying (%)	0676	10.43	0.1200	11.14	0.3859	0.1146	70	-1.85	001.99	0
001	Loss on Drying (%)	2259	10.49	0.0620	11.14	0.3859	0.1146	70	-1.70	001.07	0
001	Loss on Drying (%)	0028	10.59	0.3300	11.14	0.3859	0.1146	70	-1.44	001.07	0
001	Loss on Drying (%)	0353	10.61	0.0400	11.14	0.3859	0.1146	70	-1.38	001.07	0
001	Loss on Drying (%)	0504	10.62	0.0000	11.14	0.3859	0.1146	70	-1.35	001.00	0
001	Loss on Drying (%)	0010	10.65	0.1000	11.14	0.3859	0.1146	70	-1.28	001.07	0
001	Loss on Drying (%)	0142	10.68	0.1700	11.14	0.3859	0.1146	70	-1.21	001.07	0
001	Loss on Drying (%)	0510	10.70	0.0000	11.14	0.3859	0.1146	70	-1.15	001.99	0
001	Loss on Drying (%)	0034	10.72	0.0600	11.14	0.3859	0.1146	70	-1.09	001.07	0
001	Loss on Drying (%)	0083	10.81	0.0400	11.14	0.3859	0.1146	70	-0.86	001.07	0
001	Loss on Drying (%)	2319	10.85	0.1000	11.14	0.3859	0.1146	70	-0.76	001.99	0
001	Loss on Drying (%)	0226	10.86	0.0200	11.14	0.3859	0.1146	70	-0.73	001.07	0
001	Loss on Drying (%)	0027	10.87	0.0790	11.14	0.3859	0.1146	70	-0.71	001.00	0
001	Loss on Drying (%)	0027	10.87	0.1250	11.14	0.3859	0.1146	70	-0.71	001.07	0
001	Loss on Drying (%)	0171	10.91	0.0700	11.14	0.3859	0.1146	70	-0.61	001.07	0
001	Loss on Drying (%)	0278	10.94	0.0100	11.14	0.3859	0.1146	70	-0.54	001.07	0
001	Loss on Drying (%)	0345	10.95	0.1000	11.14	0.3859	0.1146	70	-0.50	001.07	0
001	Loss on Drying (%)	0366	10.95	0.1000	11.14	0.3859	0.1146	70	-0.50	001.07	0
001	Loss on Drying (%)	0683	10.96	0.0100	11.14	0.3859	0.1146	70	-0.49	001.07	0
001	Loss on Drying (%)	0066	10.97	0.0900	11.14	0.3859	0.1146	70	-0.46	001.07	0
001	Loss on Drying (%)	0042	10.98	0.2400	11.14	0.3859	0.1146	70	-0.42	001.99	0
001	Loss on Drying (%)	0169	11.00	0.0200	11.14	0.3859	0.1146	70	-0.37	001.00	0
001	Loss on Drying (%)	2022	11.03	0.0000	11.14	0.3859	0.1146	70	-0.29	001.07	0
001	Loss on Drying (%)	0505	11.05	0.0300	11.14	0.3859	0.1146	70	-0.25	001.07	0
001	Loss on Drying (%)	0037	11.06	0.1000	11.14	0.3859	0.1146	70	-0.21	001.99	0
001	Loss on Drying (%)	0407	11.06	0.0477	11.14	0.3859	0.1146	70	-0.20	001.07	0
001	Loss on Drying (%)	0045	11.08	0.7800	11.14	0.3859	0.1146	70	-0.16	001.07	0
001	Loss on Drying (%)	0074	11.09	0.0900	11.14	0.3859	0.1146	70	-0.15	001.07	0
001	Loss on Drying (%)	2190	11.14	0.0100	11.14	0.3859	0.1146	70	-0.02	001.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
001	Loss on Drying (%)	0590	11.14	0.0200	11.14	0.3859	0.1146	70	-0.01	001.07	0
001	Loss on Drying (%)	0644	11.15	0.0250	11.14	0.3859	0.1146	70	0.02	001.05	0
001	Loss on Drying (%)	0581	11.15	0.1000	11.14	0.3859	0.1146	70	0.02	001.07	0
001	Loss on Drying (%)	0309	11.17	0.0600	11.14	0.3859	0.1146	70	0.07	001.00	0
001	Loss on Drying (%)	0596	11.20	0.0300	11.14	0.3859	0.1146	70	0.14	001.07	0
001	Loss on Drying (%)	2009	11.20	0.0101	11.14	0.3859	0.1146	70	0.15	001.07	0
001	Loss on Drying (%)	0049	11.22	0.1700	11.14	0.3859	0.1146	70	0.19	001.07	0
001	Loss on Drying (%)	0003	11.23	0.1100	11.14	0.3859	0.1146	70	0.21	001.07	0
001	Loss on Drying (%)	0186	11.23	0.0800	11.14	0.3859	0.1146	70	0.23	001.99	0
001	Loss on Drying (%)	0148	11.24	0.1000	11.14	0.3859	0.1146	70	0.25	001.07	0
001	Loss on Drying (%)	0571	11.24	0.0200	11.14	0.3859	0.1146	70	0.25	001.07	0
001	Loss on Drying (%)	2192	11.24	0.0400	11.14	0.3859	0.1146	70	0.25	001.07	0
001	Loss on Drying (%)	0019	11.28	0.0800	11.14	0.3859	0.1146	70	0.36	001.99	0
001	Loss on Drying (%)	0638	11.30	0.0000	11.14	0.3859	0.1146	70	0.41	001.07	0
001	Loss on Drying (%)	0876	11.30	0.0000	11.14	0.3859	0.1146	70	0.41	001.07	0
001	Loss on Drying (%)	0910	11.30	0.2000	11.14	0.3859	0.1146	70	0.41	001.99	0
001	Loss on Drying (%)	0675	11.31	0.0600	11.14	0.3859	0.1146	70	0.43	001.07	0
001	Loss on Drying (%)	2246	11.32	0.0800	11.14	0.3859	0.1146	70	0.46	001.99	0
001	Loss on Drying (%)	2268	11.35	0.0000	11.14	0.3859	0.1146	70	0.54	001.99	0
001	Loss on Drying (%)	0650	11.35	0.1000	11.14	0.3859	0.1146	70	0.54	001.07	0
001	Loss on Drying (%)	2073	11.37	0.1800	11.14	0.3859	0.1146	70	0.59	001.00	0
001	Loss on Drying (%)	0098	11.44	0.2900	11.14	0.3859	0.1146	70	0.76	001.07	0
001	Loss on Drying (%)	0033	11.45	0.1800	11.14	0.3859	0.1146	70	0.80	001.07	0
001	Loss on Drying (%)	0035	11.45	0.0400	11.14	0.3859	0.1146	70	0.80	001.07	0
001	Loss on Drying (%)	2059	11.46	0.0200	11.14	0.3859	0.1146	70	0.82	001.07	0
001	Loss on Drying (%)	0918	11.49	0.0300	11.14	0.3859	0.1146	70	0.89	001.99	0
001	Loss on Drying (%)	0689	11.50	0.0000	11.14	0.3859	0.1146	70	0.93	001.07	0
001	Loss on Drying (%)	0843	11.54	0.1600	11.14	0.3859	0.1146	70	1.03	001.03	0
001	Loss on Drying (%)	0843	11.54	0.1600	11.14	0.3859	0.1146	70	1.03	001.07	0
001	Loss on Drying (%)	0643	11.55	0.0200	11.14	0.3859	0.1146	70	1.06	001.07	0
001	Loss on Drying (%)	0693	11.56	0.4620	11.14	0.3859	0.1146	70	1.08	001.07	0
001	Loss on Drying (%)	0590	11.62	0.2100	11.14	0.3859	0.1146	70	1.23	001.99	0
001	Loss on Drying (%)	0357	11.73	0.0358	11.14	0.3859	0.1146	70	1.53	001.99	0
001	Loss on Drying (%)	0183	11.80	0.0320	11.14	0.3859	0.1146	70	1.69	001.00	0
001	Loss on Drying (%)	0852	11.80	0.8000	11.14	0.3859	0.1146	70	1.70	001.99	0
001	Loss on Drying (%)	0536	11.89	0.0100	11.14	0.3859	0.1146	70	1.92	001.99	0
001	Loss on Drying (%)	0682	12.01	0.0000	11.14	0.3859	0.1146	70	2.25	001.07	0
001	Loss on Drying (%)	0297	12.13	0.0800	11.14	0.3859	0.1146	70	2.56	001.07	0
001	Loss on Drying (%)	0853	12.05	1.500	11.14	0.3859	0.1146	70	2.35	001.07	1
001	Loss on Drying (%)	2345	88.79	0.4200	11.14	0.3859	0.1146	70	201.20	001.07	2
002	Protein, Crude (%)	0618	7.611	0.0290	47.41	0.5820	0.3321	163	-68.37	002.06	0
002	Protein, Crude (%)	2073	43.45	0.1400	47.41	0.5820	0.3321	163	-6.80	002.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0553	44.05	0.1500	47.41	0.5820	0.3321	163	-5.78	002.11	0
002	Protein, Crude (%)	0581	44.85	0.1000	47.41	0.5820	0.3321	163	-4.39	002.05	0
002	Protein, Crude (%)	0536	45.15	0.0300	47.41	0.5820	0.3321	163	-3.89	002.05	0
002	Protein, Crude (%)	0852	45.25	0.3000	47.41	0.5820	0.3321	163	-3.71	002.05	0
002	Protein, Crude (%)	2200	45.90	0.2000	47.41	0.5820	0.3321	163	-2.59	002.05	0
002	Protein, Crude (%)	0098	46.06	0.2000	47.41	0.5820	0.3321	163	-2.32	002.08	0
002	Protein, Crude (%)	2246	46.19	0.1200	47.41	0.5820	0.3321	163	-2.09	002.05	0
002	Protein, Crude (%)	0689	46.25	0.5000	47.41	0.5820	0.3321	163	-1.99	002.05	0
002	Protein, Crude (%)	2259	46.38	0.1200	47.41	0.5820	0.3321	163	-1.77	002.01	0
002	Protein, Crude (%)	2199	46.45	0.7000	47.41	0.5820	0.3321	163	-1.65	002.05	0
002	Protein, Crude (%)	0028	46.50	0.4600	47.41	0.5820	0.3321	163	-1.56	002.06	0
002	Protein, Crude (%)	0596	46.51	0.3000	47.41	0.5820	0.3321	163	-1.54	002.05	0
002	Protein, Crude (%)	2322	46.53	0.0800	47.41	0.5820	0.3321	163	-1.51	002.01	0
002	Protein, Crude (%)	0910	46.55	0.9000	47.41	0.5820	0.3321	163	-1.47	002.01	0
002	Protein, Crude (%)	2006	46.57	0.4500	47.41	0.5820	0.3321	163	-1.45	002.05	0
002	Protein, Crude (%)	2097	46.58	0.4526	47.41	0.5820	0.3321	163	-1.42	002.05	0
002	Protein, Crude (%)	0563	46.58	0.1695	47.41	0.5820	0.3321	163	-1.42	002.08	0
002	Protein, Crude (%)	2009	46.59	0.0576	47.41	0.5820	0.3321	163	-1.40	002.05	0
002	Protein, Crude (%)	0674	46.60	0.0000	47.41	0.5820	0.3321	163	-1.39	002.01	0
002	Protein, Crude (%)	0650	46.60	0.2000	47.41	0.5820	0.3321	163	-1.39	002.06	0
002	Protein, Crude (%)	0968	46.66	0.0300	47.41	0.5820	0.3321	163	-1.29	002.01	0
002	Protein, Crude (%)	2193	46.66	0.2800	47.41	0.5820	0.3321	163	-1.28	002.01	0
002	Protein, Crude (%)	0964	46.68	0.3330	47.41	0.5820	0.3321	163	-1.25	002.06	0
002	Protein, Crude (%)	0683	46.71	0.2100	47.41	0.5820	0.3321	163	-1.21	002.05	0
002	Protein, Crude (%)	2022	46.71	0.0300	47.41	0.5820	0.3321	163	-1.21	002.05	0
002	Protein, Crude (%)	2181	46.81	0.1900	47.41	0.5820	0.3321	163	-1.04	002.05	0
002	Protein, Crude (%)	2337	46.82	0.0700	47.41	0.5820	0.3321	163	-1.02	002.05	0
002	Protein, Crude (%)	0652	46.95	0.1000	47.41	0.5820	0.3321	163	-0.79	002.01	0
002	Protein, Crude (%)	0948	46.96	0.0000	47.41	0.5820	0.3321	163	-0.77	002.05	0
002	Protein, Crude (%)	0504	47.00	0.0700	47.41	0.5820	0.3321	163	-0.71	002.04	0
002	Protein, Crude (%)	0638	47.00	0.0000	47.41	0.5820	0.3321	163	-0.70	002.04	0
002	Protein, Crude (%)	0345	47.00	0.2000	47.41	0.5820	0.3321	163	-0.70	002.06	0
002	Protein, Crude (%)	0723	47.01	0.1400	47.41	0.5820	0.3321	163	-0.68	002.06	0
002	Protein, Crude (%)	2303	47.01	0.1600	47.41	0.5820	0.3321	163	-0.68	002.06	0
002	Protein, Crude (%)	0939	47.02	0.3400	47.41	0.5820	0.3321	163	-0.67	002.05	0
002	Protein, Crude (%)	2188	47.03	0.0700	47.41	0.5820	0.3321	163	-0.66	002.01	0
002	Protein, Crude (%)	0870	47.08	0.4330	47.41	0.5820	0.3321	163	-0.57	002.01	0
002	Protein, Crude (%)	2184	47.09	0.0700	47.41	0.5820	0.3321	163	-0.55	002.06	0
002	Protein, Crude (%)	0859	47.09	0.0790	47.41	0.5820	0.3321	163	-0.55	002.06	0
002	Protein, Crude (%)	0870	47.09	0.1257	47.41	0.5820	0.3321	163	-0.54	002.06	0
002	Protein, Crude (%)	0941	47.10	0.3500	47.41	0.5820	0.3321	163	-0.54	002.05	0
002	Protein, Crude (%)	0953	47.10	1.100	47.41	0.5820	0.3321	163	-0.53	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0042	47.10	0.2000	47.41	0.5820	0.3321	163	-0.53	002.06	0
002	Protein, Crude (%)	0511	47.13	1.020	47.41	0.5820	0.3321	163	-0.48	002.06	0
002	Protein, Crude (%)	0354	47.14	0.0000	47.41	0.5820	0.3321	163	-0.46	002.05	0
002	Protein, Crude (%)	0164	47.16	0.3100	47.41	0.5820	0.3321	163	-0.43	002.01	0
002	Protein, Crude (%)	0653	47.16	0.0800	47.41	0.5820	0.3321	163	-0.43	002.01	0
002	Protein, Crude (%)	2268	47.16	1.520	47.41	0.5820	0.3321	163	-0.43	002.06	0
002	Protein, Crude (%)	2329	47.18	0.1000	47.41	0.5820	0.3321	163	-0.39	002.01	0
002	Protein, Crude (%)	0505	47.19	0.1500	47.41	0.5820	0.3321	163	-0.38	002.06	0
002	Protein, Crude (%)	0027	47.19	0.7330	47.41	0.5820	0.3321	163	-0.37	002.06	0
002	Protein, Crude (%)	0520	47.20	0.7300	47.41	0.5820	0.3321	163	-0.37	002.06	0
002	Protein, Crude (%)	2168	47.20	0.0100	47.41	0.5820	0.3321	163	-0.37	002.06	0
002	Protein, Crude (%)	0045	47.20	0.4000	47.41	0.5820	0.3321	163	-0.36	002.06	0
002	Protein, Crude (%)	0035	47.22	0.1000	47.41	0.5820	0.3321	163	-0.32	002.06	0
002	Protein, Crude (%)	2174	47.25	0.0300	47.41	0.5820	0.3321	163	-0.28	002.06	0
002	Protein, Crude (%)	2302	47.25	0.1200	47.41	0.5820	0.3321	163	-0.27	002.06	0
002	Protein, Crude (%)	2058	47.29	0.1100	47.41	0.5820	0.3321	163	-0.21	002.04	0
002	Protein, Crude (%)	0098	47.30	0.8000	47.41	0.5820	0.3321	163	-0.19	002.06	0
002	Protein, Crude (%)	0148	47.30	0.0800	47.41	0.5820	0.3321	163	-0.19	002.06	0
002	Protein, Crude (%)	0956	47.30	0.4000	47.41	0.5820	0.3321	163	-0.19	002.06	0
002	Protein, Crude (%)	0651	47.30	0.2473	47.41	0.5820	0.3321	163	-0.18	002.05	0
002	Protein, Crude (%)	0860	47.31	0.0000	47.41	0.5820	0.3321	163	-0.18	002.01	0
002	Protein, Crude (%)	0019	47.37	0.0500	47.41	0.5820	0.3321	163	-0.07	002.06	0
002	Protein, Crude (%)	0100	47.37	0.1900	47.41	0.5820	0.3321	163	-0.07	002.06	0
002	Protein, Crude (%)	0693	47.38	0.3750	47.41	0.5820	0.3321	163	-0.06	002.06	0
002	Protein, Crude (%)	0066	47.39	0.1800	47.41	0.5820	0.3321	163	-0.03	002.02	0
002	Protein, Crude (%)	0918	47.39	0.0000	47.41	0.5820	0.3321	163	-0.03	002.06	0
002	Protein, Crude (%)	0049	47.40	0.6100	47.41	0.5820	0.3321	163	-0.02	002.06	0
002	Protein, Crude (%)	0425	47.41	0.4300	47.41	0.5820	0.3321	163	0.00	002.06	0
002	Protein, Crude (%)	0875	47.41	0.1500	47.41	0.5820	0.3321	163	0.00	002.99	0
002	Protein, Crude (%)	0626	47.41	0.1230	47.41	0.5820	0.3321	163	0.00	002.06	0
002	Protein, Crude (%)	0006	47.41	0.0200	47.41	0.5820	0.3321	163	0.00	002.06	0
002	Protein, Crude (%)	2192	47.42	0.5300	47.41	0.5820	0.3321	163	0.01	002.05	0
002	Protein, Crude (%)	0407	47.42	0.6500	47.41	0.5820	0.3321	163	0.02	002.06	0
002	Protein, Crude (%)	0536	47.42	0.1200	47.41	0.5820	0.3321	163	0.02	002.11	0
002	Protein, Crude (%)	0297	47.43	0.2300	47.41	0.5820	0.3321	163	0.03	002.06	0
002	Protein, Crude (%)	0687	47.43	0.1500	47.41	0.5820	0.3321	163	0.03	002.06	0
002	Protein, Crude (%)	0171	47.45	0.0300	47.41	0.5820	0.3321	163	0.06	002.06	0
002	Protein, Crude (%)	0047	47.45	0.5000	47.41	0.5820	0.3321	163	0.07	002.06	0
002	Protein, Crude (%)	0354	47.45	0.0600	47.41	0.5820	0.3321	163	0.07	002.06	0
002	Protein, Crude (%)	0596	47.46	0.0900	47.41	0.5820	0.3321	163	0.08	002.06	0
002	Protein, Crude (%)	0164	47.47	0.0600	47.41	0.5820	0.3321	163	0.11	002.06	0
002	Protein, Crude (%)	0598	47.47	0.0200	47.41	0.5820	0.3321	163	0.11	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0682	47.47	0.0000	47.41	0.5820	0.3321	163	0.11	002.06	0
002	Protein, Crude (%)	0848	47.48	0.2500	47.41	0.5820	0.3321	163	0.12	002.06	0
002	Protein, Crude (%)	0539	47.49	0.1300	47.41	0.5820	0.3321	163	0.13	002.06	0
002	Protein, Crude (%)	0914	47.49	0.0100	47.41	0.5820	0.3321	163	0.13	002.06	0
002	Protein, Crude (%)	2190	47.49	0.0200	47.41	0.5820	0.3321	163	0.14	002.05	0
002	Protein, Crude (%)	0644	47.50	0.2150	47.41	0.5820	0.3321	163	0.15	002.06	0
002	Protein, Crude (%)	2345	47.50	0.2000	47.41	0.5820	0.3321	163	0.16	002.00	0
002	Protein, Crude (%)	0010	47.50	0.4000	47.41	0.5820	0.3321	163	0.16	002.06	0
002	Protein, Crude (%)	0510	47.50	0.0000	47.41	0.5820	0.3321	163	0.16	002.06	0
002	Protein, Crude (%)	0884	47.50	0.2000	47.41	0.5820	0.3321	163	0.16	002.06	0
002	Protein, Crude (%)	0948	47.52	0.0400	47.41	0.5820	0.3321	163	0.19	002.06	0
002	Protein, Crude (%)	2326	47.53	0.5600	47.41	0.5820	0.3321	163	0.21	002.06	0
002	Protein, Crude (%)	0357	47.53	0.0625	47.41	0.5820	0.3321	163	0.21	002.06	0
002	Protein, Crude (%)	0144	47.54	0.0500	47.41	0.5820	0.3321	163	0.22	002.06	0
002	Protein, Crude (%)	0553	47.55	0.1700	47.41	0.5820	0.3321	163	0.24	002.06	0
002	Protein, Crude (%)	0646	47.58	0.0000	47.41	0.5820	0.3321	163	0.30	002.06	0
002	Protein, Crude (%)	0183	47.60	0.2700	47.41	0.5820	0.3321	163	0.32	002.06	0
002	Protein, Crude (%)	0242	47.60	0.0100	47.41	0.5820	0.3321	163	0.32	002.06	0
002	Protein, Crude (%)	0142	47.60	0.0600	47.41	0.5820	0.3321	163	0.33	002.06	0
002	Protein, Crude (%)	0208	47.60	0.2000	47.41	0.5820	0.3321	163	0.33	002.06	0
002	Protein, Crude (%)	0309	47.61	0.1240	47.41	0.5820	0.3321	163	0.35	002.06	0
002	Protein, Crude (%)	0546	47.62	0.1100	47.41	0.5820	0.3321	163	0.36	002.06	0
002	Protein, Crude (%)	0353	47.62	0.0800	47.41	0.5820	0.3321	163	0.36	002.06	0
002	Protein, Crude (%)	0574	47.62	0.1800	47.41	0.5820	0.3321	163	0.36	002.06	0
002	Protein, Crude (%)	0123	47.67	0.0300	47.41	0.5820	0.3321	163	0.44	002.06	0
002	Protein, Crude (%)	0504	47.68	0.3500	47.41	0.5820	0.3321	163	0.46	002.06	0
002	Protein, Crude (%)	0590	47.68	0.3300	47.41	0.5820	0.3321	163	0.46	002.06	0
002	Protein, Crude (%)	0226	47.69	0.3800	47.41	0.5820	0.3321	163	0.48	002.06	0
002	Protein, Crude (%)	0033	47.70	0.2000	47.41	0.5820	0.3321	163	0.50	002.06	0
002	Protein, Crude (%)	0366	47.70	0.0000	47.41	0.5820	0.3321	163	0.50	002.06	0
002	Protein, Crude (%)	0910	47.70	0.2000	47.41	0.5820	0.3321	163	0.50	002.06	0
002	Protein, Crude (%)	0541	47.73	0.4400	47.41	0.5820	0.3321	163	0.55	002.06	0
002	Protein, Crude (%)	0004	47.74	0.7500	47.41	0.5820	0.3321	163	0.56	002.06	0
002	Protein, Crude (%)	0009	47.75	0.0200	47.41	0.5820	0.3321	163	0.59	002.06	0
002	Protein, Crude (%)	0263	47.75	0.1060	47.41	0.5820	0.3321	163	0.59	002.06	0
002	Protein, Crude (%)	0968	47.77	0.1800	47.41	0.5820	0.3321	163	0.62	002.06	0
002	Protein, Crude (%)	0014	47.80	0.2000	47.41	0.5820	0.3321	163	0.67	002.06	0
002	Protein, Crude (%)	0876	47.80	0.2000	47.41	0.5820	0.3321	163	0.67	002.06	0
002	Protein, Crude (%)	0675	47.81	0.0500	47.41	0.5820	0.3321	163	0.68	002.06	0
002	Protein, Crude (%)	0036	47.81	0.0520	47.41	0.5820	0.3321	163	0.69	002.06	0
002	Protein, Crude (%)	0034	47.81	0.4270	47.41	0.5820	0.3321	163	0.69	002.06	0
002	Protein, Crude (%)	0529	47.82	0.1200	47.41	0.5820	0.3321	163	0.71	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0083	47.83	0.0060	47.41	0.5820	0.3321	163	0.72	002.06	0
002	Protein, Crude (%)	0726	47.84	0.0800	47.41	0.5820	0.3321	163	0.74	002.06	0
002	Protein, Crude (%)	0300	47.85	0.7000	47.41	0.5820	0.3321	163	0.76	002.06	0
002	Protein, Crude (%)	0037	47.86	0.5400	47.41	0.5820	0.3321	163	0.78	002.06	0
002	Protein, Crude (%)	0843	47.86	0.2000	47.41	0.5820	0.3321	163	0.78	002.06	0
002	Protein, Crude (%)	0508	47.86	0.1150	47.41	0.5820	0.3321	163	0.78	002.06	0
002	Protein, Crude (%)	0108	47.89	0.0100	47.41	0.5820	0.3321	163	0.82	002.06	0
002	Protein, Crude (%)	0003	47.90	0.1200	47.41	0.5820	0.3321	163	0.85	002.06	0
002	Protein, Crude (%)	0278	47.90	0.6000	47.41	0.5820	0.3321	163	0.85	002.06	0
002	Protein, Crude (%)	0169	47.92	1.500	47.41	0.5820	0.3321	163	0.88	002.00	0
002	Protein, Crude (%)	0941	47.93	0.1700	47.41	0.5820	0.3321	163	0.89	002.06	0
002	Protein, Crude (%)	0229	47.94	0.0200	47.41	0.5820	0.3321	163	0.91	002.06	0
002	Protein, Crude (%)	0265	47.95	0.3000	47.41	0.5820	0.3321	163	0.93	002.06	0
002	Protein, Crude (%)	2319	47.95	0.1000	47.41	0.5820	0.3321	163	0.93	002.06	0
002	Protein, Crude (%)	2290	48.02	0.0100	47.41	0.5820	0.3321	163	1.04	002.11	0
002	Protein, Crude (%)	0062	48.06	0.4100	47.41	0.5820	0.3321	163	1.11	002.06	0
002	Protein, Crude (%)	0571	48.12	0.1400	47.41	0.5820	0.3321	163	1.22	002.06	0
002	Protein, Crude (%)	0011	48.14	0.0800	47.41	0.5820	0.3321	163	1.26	002.11	0
002	Protein, Crude (%)	0590	48.16	0.2000	47.41	0.5820	0.3321	163	1.29	002.11	0
002	Protein, Crude (%)	0190	48.19	0.0500	47.41	0.5820	0.3321	163	1.34	002.06	0
002	Protein, Crude (%)	0676	48.19	0.2500	47.41	0.5820	0.3321	163	1.34	002.06	0
002	Protein, Crude (%)	2089	48.27	0.0300	47.41	0.5820	0.3321	163	1.47	002.06	0
002	Protein, Crude (%)	2129	48.27	0.3010	47.41	0.5820	0.3321	163	1.47	002.06	0
002	Protein, Crude (%)	2150	48.30	0.0800	47.41	0.5820	0.3321	163	1.53	002.06	0
002	Protein, Crude (%)	2076	48.34	1.214	47.41	0.5820	0.3321	163	1.61	002.06	0
002	Protein, Crude (%)	0853	48.40	0.4000	47.41	0.5820	0.3321	163	1.70	002.06	0
002	Protein, Crude (%)	2059	48.50	0.1500	47.41	0.5820	0.3321	163	1.87	002.06	0
002	Protein, Crude (%)	2213	48.53	0.1200	47.41	0.5820	0.3321	163	1.93	002.06	0
002	Protein, Crude (%)	2105	48.72	0.1400	47.41	0.5820	0.3321	163	2.25	002.11	0
002	Protein, Crude (%)	0011	48.74	0.1700	47.41	0.5820	0.3321	163	2.28	002.06	0
002	Protein, Crude (%)	0852	48.86	0.6600	47.41	0.5820	0.3321	163	2.49	002.11	0
002	Protein, Crude (%)	0692	48.97	0.1900	47.41	0.5820	0.3321	163	2.68	002.06	0
002	Protein, Crude (%)	0685	49.95	0.2000	47.41	0.5820	0.3321	163	4.37	002.01	0
002	Protein, Crude (%)	0297	50.85	0.0100	47.41	0.5820	0.3321	163	5.91	002.11	0
002	Protein, Crude (%)	0612	57.63	15.75	47.41	0.5820	0.3321	163	17.55	002.06	0
002	Protein, Crude (%)	0202	261.7	428.7	47.41	0.5820	0.3321	163	368.08	002.06	2
003	Fat, Crude (%)	0504	0.5900	0.0400	1.004	0.2398	0.0795	119	-1.73	003.01	0
003	Fat, Crude (%)	0242	0.6000	0.0400	1.004	0.2398	0.0795	119	-1.68	003.10	0
003	Fat, Crude (%)	0042	0.6000	0.1200	1.004	0.2398	0.0795	119	-1.68	003.10	0
003	Fat, Crude (%)	0144	0.6100	0.1200	1.004	0.2398	0.0795	119	-1.64	003.14	0
003	Fat, Crude (%)	0574	0.6200	0.0600	1.004	0.2398	0.0795	119	-1.60	003.14	0
003	Fat, Crude (%)	0968	0.6200	0.2400	1.004	0.2398	0.0795	119	-1.60	003.14	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
003	Fat, Crude (%)	0300	0.6250	0.0100	1.004	0.2398	0.0795	119	-1.58	003.14	0
003	Fat, Crude (%)	0263	0.6415	0.0330	1.004	0.2398	0.0795	119	-1.51	003.10	0
003	Fat, Crude (%)	0948	0.6950	0.0900	1.004	0.2398	0.0795	119	-1.29	003.09	0
003	Fat, Crude (%)	0108	0.6950	0.2900	1.004	0.2398	0.0795	119	-1.29	003.14	0
003	Fat, Crude (%)	0098	0.7150	0.1500	1.004	0.2398	0.0795	119	-1.20	003.10	0
003	Fat, Crude (%)	2073	0.7200	0.0600	1.004	0.2398	0.0795	119	-1.18	003.10	0
003	Fat, Crude (%)	0505	0.7400	0.2400	1.004	0.2398	0.0795	119	-1.10	003.14	0
003	Fat, Crude (%)	0171	0.7500	0.1200	1.004	0.2398	0.0795	119	-1.06	003.12	0
003	Fat, Crude (%)	0036	0.7621	0.0107	1.004	0.2398	0.0795	119	-1.01	003.10	0
003	Fat, Crude (%)	0034	0.7650	0.0100	1.004	0.2398	0.0795	119	-1.00	003.14	0
003	Fat, Crude (%)	0693	0.7685	0.0170	1.004	0.2398	0.0795	119	-0.98	003.10	0
003	Fat, Crude (%)	0726	0.7750	0.0300	1.004	0.2398	0.0795	119	-0.95	003.14	0
003	Fat, Crude (%)	0675	0.7950	0.0100	1.004	0.2398	0.0795	119	-0.87	003.09	0
003	Fat, Crude (%)	0033	0.7950	0.0300	1.004	0.2398	0.0795	119	-0.87	003.13	0
003	Fat, Crude (%)	0675	0.7950	0.0100	1.004	0.2398	0.0795	119	-0.87	003.14	0
003	Fat, Crude (%)	0183	0.7955	0.3190	1.004	0.2398	0.0795	119	-0.87	003.14	0
003	Fat, Crude (%)	0278	0.8000	0.2000	1.004	0.2398	0.0795	119	-0.85	003.14	0
003	Fat, Crude (%)	0529	0.8100	0.1400	1.004	0.2398	0.0795	119	-0.81	003.14	0
003	Fat, Crude (%)	2022	0.8250	0.0100	1.004	0.2398	0.0795	119	-0.75	003.10	0
003	Fat, Crude (%)	2076	0.8270	0.1301	1.004	0.2398	0.0795	119	-0.74	003.14	0
003	Fat, Crude (%)	0674	0.8300	0.0000	1.004	0.2398	0.0795	119	-0.73	003.06	0
003	Fat, Crude (%)	0848	0.8350	0.0300	1.004	0.2398	0.0795	119	-0.70	003.14	0
003	Fat, Crude (%)	0011	0.8400	0.0200	1.004	0.2398	0.0795	119	-0.68	003.11	0
003	Fat, Crude (%)	0074	0.8400	0.0800	1.004	0.2398	0.0795	119	-0.68	003.14	0
003	Fat, Crude (%)	0598	0.8450	0.0300	1.004	0.2398	0.0795	119	-0.66	003.14	0
003	Fat, Crude (%)	0229	0.8500	0.2400	1.004	0.2398	0.0795	119	-0.64	003.14	0
003	Fat, Crude (%)	0581	0.8500	0.1000	1.004	0.2398	0.0795	119	-0.64	003.14	0
003	Fat, Crude (%)	0876	0.8500	0.1000	1.004	0.2398	0.0795	119	-0.64	003.14	0
003	Fat, Crude (%)	0956	0.8500	0.1000	1.004	0.2398	0.0795	119	-0.64	003.14	0
003	Fat, Crude (%)	0098	0.8550	0.0500	1.004	0.2398	0.0795	119	-0.62	003.13	0
003	Fat, Crude (%)	0510	0.8650	0.0100	1.004	0.2398	0.0795	119	-0.58	003.14	0
003	Fat, Crude (%)	0226	0.8850	0.0900	1.004	0.2398	0.0795	119	-0.50	003.14	0
003	Fat, Crude (%)	0004	0.8900	0.0200	1.004	0.2398	0.0795	119	-0.48	003.09	0
003	Fat, Crude (%)	0265	0.8900	0.0200	1.004	0.2398	0.0795	119	-0.48	003.14	0
003	Fat, Crude (%)	0004	0.8950	0.0100	1.004	0.2398	0.0795	119	-0.45	003.14	0
003	Fat, Crude (%)	0345	0.9000	0.1000	1.004	0.2398	0.0795	119	-0.43	003.00	0
003	Fat, Crude (%)	0638	0.9000	0.0000	1.004	0.2398	0.0795	119	-0.43	003.09	0
003	Fat, Crude (%)	0357	0.9000	0.0200	1.004	0.2398	0.0795	119	-0.43	003.14	0
003	Fat, Crude (%)	0948	0.9000	0.0200	1.004	0.2398	0.0795	119	-0.43	003.14	0
003	Fat, Crude (%)	2105	0.9050	0.1100	1.004	0.2398	0.0795	119	-0.41	003.11	0
003	Fat, Crude (%)	0202	0.9050	0.1900	1.004	0.2398	0.0795	119	-0.41	003.14	0
003	Fat, Crude (%)	0009	0.9100	0.0600	1.004	0.2398	0.0795	119	-0.39	003.14	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
003	Fat, Crude (%)	0019	0.9100	0.0200	1.004	0.2398	0.0795	119	-0.39	003.14	0
003	Fat, Crude (%)	0011	0.9150	0.0900	1.004	0.2398	0.0795	119	-0.37	003.00	0
003	Fat, Crude (%)	0098	0.9150	0.0300	1.004	0.2398	0.0795	119	-0.37	003.09	0
003	Fat, Crude (%)	0100	0.9200	0.1800	1.004	0.2398	0.0795	119	-0.35	003.10	0
003	Fat, Crude (%)	0870	0.9290	0.0534	1.004	0.2398	0.0795	119	-0.31	003.10	0
003	Fat, Crude (%)	0083	0.9350	0.0300	1.004	0.2398	0.0795	119	-0.29	003.14	0
003	Fat, Crude (%)	0208	0.9365	0.1870	1.004	0.2398	0.0795	119	-0.28	003.13	0
003	Fat, Crude (%)	0229	0.9700	0.3600	1.004	0.2398	0.0795	119	-0.14	003.06	0
003	Fat, Crude (%)	0354	0.9900	0.0000	1.004	0.2398	0.0795	119	-0.06	003.09	0
003	Fat, Crude (%)	0297	1.000	0.0600	1.004	0.2398	0.0795	119	-0.02	003.06	0
003	Fat, Crude (%)	0407	1.001	0.1336	1.004	0.2398	0.0795	119	-0.01	003.14	0
003	Fat, Crude (%)	0148	1.015	0.0500	1.004	0.2398	0.0795	119	0.05	003.14	0
003	Fat, Crude (%)	0676	1.020	0.2000	1.004	0.2398	0.0795	119	0.07	003.10	0
003	Fat, Crude (%)	2193	1.020	0.0400	1.004	0.2398	0.0795	119	0.07	003.10	0
003	Fat, Crude (%)	2337	1.025	0.0100	1.004	0.2398	0.0795	119	0.09	003.00	0
003	Fat, Crude (%)	0045	1.025	0.0100	1.004	0.2398	0.0795	119	0.09	003.10	0
003	Fat, Crude (%)	0541	1.025	0.0500	1.004	0.2398	0.0795	119	0.09	003.10	0
003	Fat, Crude (%)	0651	1.036	0.1519	1.004	0.2398	0.0795	119	0.13	003.06	0
003	Fat, Crude (%)	0860	1.043	0.0000	1.004	0.2398	0.0795	119	0.16	003.14	0
003	Fat, Crude (%)	2009	1.044	0.0093	1.004	0.2398	0.0795	119	0.17	003.10	0
003	Fat, Crude (%)	0653	1.055	0.0100	1.004	0.2398	0.0795	119	0.21	003.00	0
003	Fat, Crude (%)	2181	1.055	0.0500	1.004	0.2398	0.0795	119	0.21	003.06	0
003	Fat, Crude (%)	0148	1.055	0.0700	1.004	0.2398	0.0795	119	0.21	003.10	0
003	Fat, Crude (%)	2129	1.060	0.0000	1.004	0.2398	0.0795	119	0.23	003.14	0
003	Fat, Crude (%)	0590	1.075	0.1100	1.004	0.2398	0.0795	119	0.30	003.09	0
003	Fat, Crude (%)	0918	1.075	0.0100	1.004	0.2398	0.0795	119	0.30	003.06	0
003	Fat, Crude (%)	0190	1.075	0.0100	1.004	0.2398	0.0795	119	0.30	003.14	0
003	Fat, Crude (%)	2190	1.085	0.0100	1.004	0.2398	0.0795	119	0.34	003.00	0
003	Fat, Crude (%)	2302	1.085	0.0900	1.004	0.2398	0.0795	119	0.34	003.14	0
003	Fat, Crude (%)	0689	1.100	0.0000	1.004	0.2398	0.0795	119	0.40	003.06	0
003	Fat, Crude (%)	0366	1.100	0.2000	1.004	0.2398	0.0795	119	0.40	003.10	0
003	Fat, Crude (%)	0689	1.100	0.0000	1.004	0.2398	0.0795	119	0.40	003.10	0
003	Fat, Crude (%)	2097	1.101	0.0479	1.004	0.2398	0.0795	119	0.40	003.06	0
003	Fat, Crude (%)	0590	1.110	0.0600	1.004	0.2398	0.0795	119	0.44	003.11	0
003	Fat, Crude (%)	0353	1.115	0.0500	1.004	0.2398	0.0795	119	0.46	003.09	0
003	Fat, Crude (%)	0682	1.120	0.0000	1.004	0.2398	0.0795	119	0.48	003.06	0
003	Fat, Crude (%)	2192	1.130	0.1200	1.004	0.2398	0.0795	119	0.53	003.00	0
003	Fat, Crude (%)	0164	1.140	0.0400	1.004	0.2398	0.0795	119	0.57	003.01	0
003	Fat, Crude (%)	0687	1.150	0.1000	1.004	0.2398	0.0795	119	0.61	003.06	0
003	Fat, Crude (%)	0884	1.150	0.1000	1.004	0.2398	0.0795	119	0.61	003.06	0
003	Fat, Crude (%)	0853	1.150	0.1000	1.004	0.2398	0.0795	119	0.61	003.14	0
003	Fat, Crude (%)	0644	1.152	0.0610	1.004	0.2398	0.0795	119	0.62	003.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
003	Fat, Crude (%)	2246	1.155	0.0100	1.004	0.2398	0.0795	119	0.63	003.13	0
003	Fat, Crude (%)	0003	1.165	0.4700	1.004	0.2398	0.0795	119	0.67	003.14	0
003	Fat, Crude (%)	0425	1.180	0.0400	1.004	0.2398	0.0795	119	0.73	003.06	0
003	Fat, Crude (%)	0123	1.185	0.0100	1.004	0.2398	0.0795	119	0.76	003.13	0
003	Fat, Crude (%)	0066	1.190	0.1630	1.004	0.2398	0.0795	119	0.77	003.14	0
003	Fat, Crude (%)	0563	1.196	0.0097	1.004	0.2398	0.0795	119	0.80	003.01	0
003	Fat, Crude (%)	0563	1.196	0.0097	1.004	0.2398	0.0795	119	0.80	003.14	0
003	Fat, Crude (%)	2184	1.220	0.0200	1.004	0.2398	0.0795	119	0.90	003.10	0
003	Fat, Crude (%)	0169	1.225	0.0500	1.004	0.2398	0.0795	119	0.92	003.06	0
003	Fat, Crude (%)	2319	1.250	0.1000	1.004	0.2398	0.0795	119	1.03	003.10	0
003	Fat, Crude (%)	0297	1.250	0.1800	1.004	0.2398	0.0795	119	1.03	003.11	0
003	Fat, Crude (%)	0685	1.270	0.0600	1.004	0.2398	0.0795	119	1.11	003.09	0
003	Fat, Crude (%)	0536	1.270	0.0200	1.004	0.2398	0.0795	119	1.11	003.99	0
003	Fat, Crude (%)	0646	1.275	0.0500	1.004	0.2398	0.0795	119	1.13	003.13	0
003	Fat, Crude (%)	0618	1.300	0.0101	1.004	0.2398	0.0795	119	1.24	003.10	0
003	Fat, Crude (%)	2326	1.340	0.2000	1.004	0.2398	0.0795	119	1.40	003.12	0
003	Fat, Crude (%)	0520	1.355	0.0500	1.004	0.2398	0.0795	119	1.46	003.14	0
003	Fat, Crude (%)	0682	1.360	0.0000	1.004	0.2398	0.0795	119	1.48	003.14	0
003	Fat, Crude (%)	2006	1.375	0.1900	1.004	0.2398	0.0795	119	1.55	003.13	0
003	Fat, Crude (%)	0309	1.400	0.0000	1.004	0.2398	0.0795	119	1.65	003.00	0
003	Fat, Crude (%)	2089	1.430	0.2800	1.004	0.2398	0.0795	119	1.78	003.10	0
003	Fat, Crude (%)	2199	1.490	0.0400	1.004	0.2398	0.0795	119	2.03	003.12	0
003	Fat, Crude (%)	0049	1.515	0.4100	1.004	0.2398	0.0795	119	2.13	003.14	0
003	Fat, Crude (%)	0553	1.605	0.0300	1.004	0.2398	0.0795	119	2.51	003.11	0
003	Fat, Crude (%)	2188	1.630	0.0000	1.004	0.2398	0.0795	119	2.61	003.10	0
003	Fat, Crude (%)	0852	1.705	0.0300	1.004	0.2398	0.0795	119	2.92	003.11	0
003	Fat, Crude (%)	2329	1.750	0.0000	1.004	0.2398	0.0795	119	3.11	003.00	0
003	Fat, Crude (%)	0546	1.875	0.0500	1.004	0.2398	0.0795	119	3.63	003.99	0
003	Fat, Crude (%)	0941	2.050	0.1400	1.004	0.2398	0.0795	119	4.36	003.99	0
003	Fat, Crude (%)	0037	0.6300	0.5600	1.004	0.2398	0.0795	119	-1.56	003.14	1
003	Fat, Crude (%)	0692	0.9000	0.6000	1.004	0.2398	0.0795	119	-0.43	003.14	1
003	Fat, Crude (%)	0910	2.475	0.6100	1.004	0.2398	0.0795	119	6.13	003.06	1
003	Fat, Crude (%)	2200	2.850	0.9000	1.004	0.2398	0.0795	119	7.70	003.12	1
003	Fat, Crude (%)	0511	0.0250	0.0500	1.004	0.2398	0.0795	119		003.14	4
004	Fiber, Crude (%)	0852	1.915	0.1300	3.502	0.3538	0.1456	108	-4.49	004.11	0
004	Fiber, Crude (%)	0011	2.335	0.0700	3.502	0.3538	0.1456	108	-3.30	004.11	0
004	Fiber, Crude (%)	0202	2.465	0.4700	3.502	0.3538	0.1456	108	-2.93	004.07	0
004	Fiber, Crude (%)	0265	2.720	0.1600	3.502	0.3538	0.1456	108	-2.21	004.07	0
004	Fiber, Crude (%)	0505	2.725	0.0900	3.502	0.3538	0.1456	108	-2.20	004.07	0
004	Fiber, Crude (%)	0144	2.740	0.2000	3.502	0.3538	0.1456	108	-2.15	004.07	0
004	Fiber, Crude (%)	0644	2.789	0.0580	3.502	0.3538	0.1456	108	-2.01	004.07	0
004	Fiber, Crude (%)	2319	2.850	0.1000	3.502	0.3538	0.1456	108	-1.84	004.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
004	Fiber, Crude (%)	2193	2.870	0.0000	3.502	0.3538	0.1456	108	-1.79	004.07	0
004	Fiber, Crude (%)	0510	3.000	0.4000	3.502	0.3538	0.1456	108	-1.42	004.07	0
004	Fiber, Crude (%)	0674	3.050	0.0000	3.502	0.3538	0.1456	108	-1.28	004.06	0
004	Fiber, Crude (%)	0229	3.050	0.1200	3.502	0.3538	0.1456	108	-1.28	004.07	0
004	Fiber, Crude (%)	0407	3.071	0.2733	3.502	0.3538	0.1456	108	-1.22	004.07	0
004	Fiber, Crude (%)	2246	3.075	0.1100	3.502	0.3538	0.1456	108	-1.21	004.03	0
004	Fiber, Crude (%)	0914	3.100	0.2000	3.502	0.3538	0.1456	108	-1.14	004.07	0
004	Fiber, Crude (%)	0918	3.145	0.1900	3.502	0.3538	0.1456	108	-1.01	004.06	0
004	Fiber, Crude (%)	0353	3.150	0.3000	3.502	0.3538	0.1456	108	-0.99	004.03	0
004	Fiber, Crude (%)	0968	3.175	0.0300	3.502	0.3538	0.1456	108	-0.92	004.06	0
004	Fiber, Crude (%)	0708	3.180	0.0200	3.502	0.3538	0.1456	108	-0.91	004.07	0
004	Fiber, Crude (%)	0169	3.200	0.0000	3.502	0.3538	0.1456	108	-0.85	004.00	0
004	Fiber, Crude (%)	0183	3.201	0.1390	3.502	0.3538	0.1456	108	-0.85	004.00	0
004	Fiber, Crude (%)	2181	3.205	0.1700	3.502	0.3538	0.1456	108	-0.84	004.07	0
004	Fiber, Crude (%)	2009	3.236	0.0670	3.502	0.3538	0.1456	108	-0.75	004.07	0
004	Fiber, Crude (%)	0083	3.240	0.0200	3.502	0.3538	0.1456	108	-0.74	004.07	0
004	Fiber, Crude (%)	0066	3.275	0.1300	3.502	0.3538	0.1456	108	-0.64	004.07	0
004	Fiber, Crude (%)	0035	3.280	0.0400	3.502	0.3538	0.1456	108	-0.63	004.07	0
004	Fiber, Crude (%)	0683	3.280	0.0200	3.502	0.3538	0.1456	108	-0.63	004.07	0
004	Fiber, Crude (%)	0520	3.290	0.0200	3.502	0.3538	0.1456	108	-0.60	004.07	0
004	Fiber, Crude (%)	2337	3.295	0.0500	3.502	0.3538	0.1456	108	-0.58	004.07	0
004	Fiber, Crude (%)	0164	3.300	0.0000	3.502	0.3538	0.1456	108	-0.57	004.00	0
004	Fiber, Crude (%)	0028	3.300	0.0000	3.502	0.3538	0.1456	108	-0.57	004.07	0
004	Fiber, Crude (%)	2006	3.315	0.1900	3.502	0.3538	0.1456	108	-0.53	004.06	0
004	Fiber, Crude (%)	2076	3.335	0.7790	3.502	0.3538	0.1456	108	-0.47	004.00	0
004	Fiber, Crude (%)	0004	3.340	0.4200	3.502	0.3538	0.1456	108	-0.46	004.07	0
004	Fiber, Crude (%)	0309	3.341	0.0371	3.502	0.3538	0.1456	108	-0.45	004.00	0
004	Fiber, Crude (%)	0123	3.345	0.0500	3.502	0.3538	0.1456	108	-0.44	004.07	0
004	Fiber, Crude (%)	0278	3.350	0.1000	3.502	0.3538	0.1456	108	-0.43	004.07	0
004	Fiber, Crude (%)	0884	3.350	0.5000	3.502	0.3538	0.1456	108	-0.43	004.00	0
004	Fiber, Crude (%)	0511	3.350	0.3000	3.502	0.3538	0.1456	108	-0.43	004.07	0
004	Fiber, Crude (%)	0653	3.365	0.0100	3.502	0.3538	0.1456	108	-0.39	004.07	0
004	Fiber, Crude (%)	0643	3.372	0.0310	3.502	0.3538	0.1456	108	-0.37	004.07	0
004	Fiber, Crude (%)	0009	3.390	0.1200	3.502	0.3538	0.1456	108	-0.32	004.07	0
004	Fiber, Crude (%)	2200	3.400	0.0000	3.502	0.3538	0.1456	108	-0.29	004.06	0
004	Fiber, Crude (%)	0366	3.400	0.2000	3.502	0.3538	0.1456	108	-0.29	004.07	0
004	Fiber, Crude (%)	0596	3.400	0.1200	3.502	0.3538	0.1456	108	-0.29	004.07	0
004	Fiber, Crude (%)	0300	3.400	0.1800	3.502	0.3538	0.1456	108	-0.29	004.07	0
004	Fiber, Crude (%)	2259	3.410	0.0660	3.502	0.3538	0.1456	108	-0.26	004.06	0
004	Fiber, Crude (%)	0263	3.469	0.0020	3.502	0.3538	0.1456	108	-0.09	004.07	0
004	Fiber, Crude (%)	0860	3.480	0.0000	3.502	0.3538	0.1456	108	-0.06	004.07	0
004	Fiber, Crude (%)	0948	3.495	0.1900	3.502	0.3538	0.1456	108	-0.02	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
004	Fiber, Crude (%)	0876	3.500	0.4000	3.502	0.3538	0.1456	108	-0.01	004.07	0
004	Fiber, Crude (%)	2290	3.500	0.2000	3.502	0.3538	0.1456	108	-0.01	004.11	0
004	Fiber, Crude (%)	0003	3.515	0.0300	3.502	0.3538	0.1456	108	0.04	004.07	0
004	Fiber, Crude (%)	0964	3.520	0.1200	3.502	0.3538	0.1456	108	0.05	004.06	0
004	Fiber, Crude (%)	0948	3.525	0.0100	3.502	0.3538	0.1456	108	0.07	004.06	0
004	Fiber, Crude (%)	0242	3.535	0.0500	3.502	0.3538	0.1456	108	0.09	004.07	0
004	Fiber, Crude (%)	0027	3.538	0.0020	3.502	0.3538	0.1456	108	0.10	004.06	0
004	Fiber, Crude (%)	0098	3.550	0.5000	3.502	0.3538	0.1456	108	0.14	004.07	0
004	Fiber, Crude (%)	0581	3.550	0.1000	3.502	0.3538	0.1456	108	0.14	004.07	0
004	Fiber, Crude (%)	0504	3.560	0.4400	3.502	0.3538	0.1456	108	0.16	004.00	0
004	Fiber, Crude (%)	0354	3.565	0.0100	3.502	0.3538	0.1456	108	0.18	004.07	0
004	Fiber, Crude (%)	0693	3.570	0.1030	3.502	0.3538	0.1456	108	0.19	004.07	0
004	Fiber, Crude (%)	0626	3.580	0.0800	3.502	0.3538	0.1456	108	0.22	004.07	0
004	Fiber, Crude (%)	0529	3.595	0.0700	3.502	0.3538	0.1456	108	0.26	004.07	0
004	Fiber, Crude (%)	0689	3.600	0.4000	3.502	0.3538	0.1456	108	0.28	004.06	0
004	Fiber, Crude (%)	0939	3.605	0.2100	3.502	0.3538	0.1456	108	0.29	004.07	0
004	Fiber, Crude (%)	0870	3.606	0.1303	3.502	0.3538	0.1456	108	0.30	004.07	0
004	Fiber, Crude (%)	0019	3.615	0.0300	3.502	0.3538	0.1456	108	0.32	004.07	0
004	Fiber, Crude (%)	0098	3.620	0.2800	3.502	0.3538	0.1456	108	0.33	004.06	0
004	Fiber, Crude (%)	0042	3.620	0.1600	3.502	0.3538	0.1456	108	0.33	004.07	0
004	Fiber, Crude (%)	0968	3.620	0.0200	3.502	0.3538	0.1456	108	0.33	004.07	0
004	Fiber, Crude (%)	0589	3.635	0.5300	3.502	0.3538	0.1456	108	0.38	004.07	0
004	Fiber, Crude (%)	0425	3.650	0.3000	3.502	0.3538	0.1456	108	0.42	004.00	0
004	Fiber, Crude (%)	0345	3.650	0.1000	3.502	0.3538	0.1456	108	0.42	004.00	0
004	Fiber, Crude (%)	0011	3.650	0.1000	3.502	0.3538	0.1456	108	0.42	004.07	0
004	Fiber, Crude (%)	0190	3.650	0.1000	3.502	0.3538	0.1456	108	0.42	004.07	0
004	Fiber, Crude (%)	0875	3.665	0.1700	3.502	0.3538	0.1456	108	0.46	004.06	0
004	Fiber, Crude (%)	2058	3.670	0.0200	3.502	0.3538	0.1456	108	0.48	004.03	0
004	Fiber, Crude (%)	0226	3.690	0.0600	3.502	0.3538	0.1456	108	0.53	004.07	0
004	Fiber, Crude (%)	0536	3.695	0.0100	3.502	0.3538	0.1456	108	0.55	004.07	0
004	Fiber, Crude (%)	0100	3.725	0.0500	3.502	0.3538	0.1456	108	0.63	004.07	0
004	Fiber, Crude (%)	2089	3.735	0.1900	3.502	0.3538	0.1456	108	0.66	004.03	0
004	Fiber, Crude (%)	2188	3.740	0.0000	3.502	0.3538	0.1456	108	0.67	004.07	0
004	Fiber, Crude (%)	0553	3.745	0.1100	3.502	0.3538	0.1456	108	0.69	004.11	0
004	Fiber, Crude (%)	0692	3.750	0.1000	3.502	0.3538	0.1456	108	0.70	004.07	0
004	Fiber, Crude (%)	0553	3.760	0.0200	3.502	0.3538	0.1456	108	0.73	004.07	0
004	Fiber, Crude (%)	0848	3.790	0.1400	3.502	0.3538	0.1456	108	0.81	004.07	0
004	Fiber, Crude (%)	0676	3.795	0.2500	3.502	0.3538	0.1456	108	0.83	004.06	0
004	Fiber, Crude (%)	0682	3.800	0.0000	3.502	0.3538	0.1456	108	0.84	004.07	0
004	Fiber, Crude (%)	0675	3.825	0.2100	3.502	0.3538	0.1456	108	0.91	004.07	0
004	Fiber, Crude (%)	0726	3.825	0.0100	3.502	0.3538	0.1456	108	0.91	004.07	0
004	Fiber, Crude (%)	0096	3.865	0.0100	3.502	0.3538	0.1456	108	1.03	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
004	Fiber, Crude (%)	2302	3.865	0.0500	3.502	0.3538	0.1456	108	1.03	004.07	0
004	Fiber, Crude (%)	0171	3.885	0.0100	3.502	0.3538	0.1456	108	1.08	004.00	0
004	Fiber, Crude (%)	0563	3.905	0.1262	3.502	0.3538	0.1456	108	1.14	004.07	0
004	Fiber, Crude (%)	0034	3.920	0.0600	3.502	0.3538	0.1456	108	1.18	004.07	0
004	Fiber, Crude (%)	0045	3.940	0.0000	3.502	0.3538	0.1456	108	1.24	004.07	0
004	Fiber, Crude (%)	0074	4.020	0.0000	3.502	0.3538	0.1456	108	1.46	004.07	0
004	Fiber, Crude (%)	2199	4.050	0.7000	3.502	0.3538	0.1456	108	1.55	004.06	0
004	Fiber, Crude (%)	2326	4.220	0.2400	3.502	0.3538	0.1456	108	2.03	004.00	0
004	Fiber, Crude (%)	0598	4.245	0.0700	3.502	0.3538	0.1456	108	2.10	004.07	0
004	Fiber, Crude (%)	2192	4.260	0.0800	3.502	0.3538	0.1456	108	2.14	004.03	0
004	Fiber, Crude (%)	0646	4.260	0.1400	3.502	0.3538	0.1456	108	2.14	004.07	0
004	Fiber, Crude (%)	2345	4.330	0.0800	3.502	0.3538	0.1456	108	2.34	004.07	0
004	Fiber, Crude (%)	2105	4.355	0.1100	3.502	0.3538	0.1456	108	2.41	004.11	0
004	Fiber, Crude (%)	0208	4.375	0.8100	3.502	0.3538	0.1456	108	2.47	004.00	0
004	Fiber, Crude (%)	2073	4.745	0.5500	3.502	0.3538	0.1456	108	3.51	004.00	0
004	Fiber, Crude (%)	0638	4.900	0.0000	3.502	0.3538	0.1456	108	3.95	004.06	0
004	Fiber, Crude (%)	0956	2.600	1.400	3.502	0.3538	0.1456	108	-2.55	004.07	1
005	Ash (%)	0511	5.450	0.3000	6.125	0.3238	0.1172	130	-2.09	005.00	0
005	Ash (%)	0062	5.470	0.3600	6.125	0.3238	0.1172	130	-2.02	005.00	0
005	Ash (%)	0169	5.480	0.0000	6.125	0.3238	0.1172	130	-1.99	005.00	0
005	Ash (%)	0953	5.570	0.1600	6.125	0.3238	0.1172	130	-1.71	005.00	0
005	Ash (%)	2268	5.645	0.0300	6.125	0.3238	0.1172	130	-1.48	005.00	0
005	Ash (%)	0004	5.690	0.0600	6.125	0.3238	0.1172	130	-1.34	005.00	0
005	Ash (%)	0968	5.695	0.0100	6.125	0.3238	0.1172	130	-1.33	005.00	0
005	Ash (%)	0651	5.723	0.0464	6.125	0.3238	0.1172	130	-1.24	005.00	0
005	Ash (%)	2150	5.725	0.2500	6.125	0.3238	0.1172	130	-1.24	005.00	0
005	Ash (%)	0035	5.755	0.0300	6.125	0.3238	0.1172	130	-1.14	005.00	0
005	Ash (%)	0142	5.770	0.0200	6.125	0.3238	0.1172	130	-1.10	005.00	0
005	Ash (%)	0589	5.785	0.1300	6.125	0.3238	0.1172	130	-1.05	005.00	0
005	Ash (%)	0852	5.785	0.0100	6.125	0.3238	0.1172	130	-1.05	005.11	0
005	Ash (%)	0229	5.795	0.4100	6.125	0.3238	0.1172	130	-1.02	005.00	0
005	Ash (%)	2329	5.795	0.0300	6.125	0.3238	0.1172	130	-1.02	005.99	0
005	Ash (%)	0652	5.800	0.2000	6.125	0.3238	0.1172	130	-1.00	005.99	0
005	Ash (%)	0505	5.805	0.0100	6.125	0.3238	0.1172	130	-0.99	005.00	0
005	Ash (%)	0539	5.810	0.1800	6.125	0.3238	0.1172	130	-0.97	005.00	0
005	Ash (%)	0045	5.815	0.1900	6.125	0.3238	0.1172	130	-0.96	005.00	0
005	Ash (%)	0848	5.815	0.0300	6.125	0.3238	0.1172	130	-0.96	005.00	0
005	Ash (%)	0726	5.815	0.1500	6.125	0.3238	0.1172	130	-0.96	005.05	0
005	Ash (%)	0870	5.830	0.0407	6.125	0.3238	0.1172	130	-0.91	005.00	0
005	Ash (%)	2337	5.835	0.0300	6.125	0.3238	0.1172	130	-0.90	005.00	0
005	Ash (%)	0034	5.837	0.0630	6.125	0.3238	0.1172	130	-0.89	005.05	0
005	Ash (%)	0644	5.839	0.1240	6.125	0.3238	0.1172	130	-0.88	005.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
005	Ash (%)	0581	5.850	0.1000	6.125	0.3238	0.1172	130	-0.85	005.00	0
005	Ash (%)	2199	5.850	0.3000	6.125	0.3238	0.1172	130	-0.85	005.00	0
005	Ash (%)	0618	5.868	0.0087	6.125	0.3238	0.1172	130	-0.79	005.00	0
005	Ash (%)	2326	5.875	0.2900	6.125	0.3238	0.1172	130	-0.77	005.00	0
005	Ash (%)	0590	5.890	0.0000	6.125	0.3238	0.1172	130	-0.73	005.05	0
005	Ash (%)	0590	5.890	0.0400	6.125	0.3238	0.1172	130	-0.73	005.11	0
005	Ash (%)	0674	5.900	0.0000	6.125	0.3238	0.1172	130	-0.70	005.00	0
005	Ash (%)	0956	5.900	0.2000	6.125	0.3238	0.1172	130	-0.70	005.00	0
005	Ash (%)	0144	5.915	0.3500	6.125	0.3238	0.1172	130	-0.65	005.00	0
005	Ash (%)	0650	5.915	0.0300	6.125	0.3238	0.1172	130	-0.65	005.00	0
005	Ash (%)	0183	5.920	0.1100	6.125	0.3238	0.1172	130	-0.63	005.00	0
005	Ash (%)	0553	5.925	0.0900	6.125	0.3238	0.1172	130	-0.62	005.00	0
005	Ash (%)	2022	5.935	0.0900	6.125	0.3238	0.1172	130	-0.59	005.00	0
005	Ash (%)	0918	5.940	0.0000	6.125	0.3238	0.1172	130	-0.57	005.00	0
005	Ash (%)	0049	5.945	0.3500	6.125	0.3238	0.1172	130	-0.56	005.00	0
005	Ash (%)	0098	5.950	0.1000	6.125	0.3238	0.1172	130	-0.54	005.00	0
005	Ash (%)	0876	5.950	0.1000	6.125	0.3238	0.1172	130	-0.54	005.00	0
005	Ash (%)	2190	5.950	0.1000	6.125	0.3238	0.1172	130	-0.54	005.05	0
005	Ash (%)	0297	5.970	0.0400	6.125	0.3238	0.1172	130	-0.48	005.00	0
005	Ash (%)	0226	5.990	0.0600	6.125	0.3238	0.1172	130	-0.42	005.00	0
005	Ash (%)	0676	5.990	0.0400	6.125	0.3238	0.1172	130	-0.42	005.99	0
005	Ash (%)	0689	5.995	0.1300	6.125	0.3238	0.1172	130	-0.40	005.05	0
005	Ash (%)	0042	6.000	0.0400	6.125	0.3238	0.1172	130	-0.39	005.00	0
005	Ash (%)	0638	6.000	0.0000	6.125	0.3238	0.1172	130	-0.39	005.05	0
005	Ash (%)	2055	6.000	0.0000	6.125	0.3238	0.1172	130	-0.39	005.99	0
005	Ash (%)	0644	6.005	0.0450	6.125	0.3238	0.1172	130	-0.37	005.02	0
005	Ash (%)	2345	6.015	0.1500	6.125	0.3238	0.1172	130	-0.34	005.05	0
005	Ash (%)	0598	6.020	0.0200	6.125	0.3238	0.1172	130	-0.33	005.00	0
005	Ash (%)	0504	6.025	0.0300	6.125	0.3238	0.1172	130	-0.31	005.00	0
005	Ash (%)	0263	6.034	0.0090	6.125	0.3238	0.1172	130	-0.28	005.05	0
005	Ash (%)	0683	6.035	0.2700	6.125	0.3238	0.1172	130	-0.28	005.05	0
005	Ash (%)	2089	6.045	0.0700	6.125	0.3238	0.1172	130	-0.25	005.00	0
005	Ash (%)	2302	6.045	0.0500	6.125	0.3238	0.1172	130	-0.25	005.00	0
005	Ash (%)	0407	6.057	0.0235	6.125	0.3238	0.1172	130	-0.21	005.00	0
005	Ash (%)	0693	6.058	0.0410	6.125	0.3238	0.1172	130	-0.21	005.00	0
005	Ash (%)	0066	6.065	0.0900	6.125	0.3238	0.1172	130	-0.19	005.05	0
005	Ash (%)	2006	6.075	0.0500	6.125	0.3238	0.1172	130	-0.16	005.00	0
005	Ash (%)	0148	6.080	0.0200	6.125	0.3238	0.1172	130	-0.14	005.00	0
005	Ash (%)	0171	6.080	0.0800	6.125	0.3238	0.1172	130	-0.14	005.00	0
005	Ash (%)	0354	6.080	0.0200	6.125	0.3238	0.1172	130	-0.14	005.00	0
005	Ash (%)	0914	6.080	0.1600	6.125	0.3238	0.1172	130	-0.14	005.00	0
005	Ash (%)	2193	6.080	0.0000	6.125	0.3238	0.1172	130	-0.14	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
005	Ash (%)	2246	6.095	0.0700	6.125	0.3238	0.1172	130	-0.09	005.05	0
005	Ash (%)	0563	6.101	0.0312	6.125	0.3238	0.1172	130	-0.08	005.00	0
005	Ash (%)	0520	6.105	0.0700	6.125	0.3238	0.1172	130	-0.06	005.00	0
005	Ash (%)	0643	6.105	0.0700	6.125	0.3238	0.1172	130	-0.06	005.00	0
005	Ash (%)	0425	6.110	0.0400	6.125	0.3238	0.1172	130	-0.05	005.00	0
005	Ash (%)	0675	6.115	0.1300	6.125	0.3238	0.1172	130	-0.03	005.00	0
005	Ash (%)	0242	6.125	0.0300	6.125	0.3238	0.1172	130	0.00	005.00	0
005	Ash (%)	0027	6.134	0.0250	6.125	0.3238	0.1172	130	0.03	005.00	0
005	Ash (%)	0345	6.135	0.0700	6.125	0.3238	0.1172	130	0.03	005.00	0
005	Ash (%)	0875	6.145	0.0300	6.125	0.3238	0.1172	130	0.06	005.99	0
005	Ash (%)	0047	6.150	0.3000	6.125	0.3238	0.1172	130	0.08	005.00	0
005	Ash (%)	0910	6.150	0.7000	6.125	0.3238	0.1172	130	0.08	005.99	0
005	Ash (%)	0278	6.170	0.2000	6.125	0.3238	0.1172	130	0.14	005.00	0
005	Ash (%)	0546	6.170	0.0000	6.125	0.3238	0.1172	130	0.14	005.99	0
005	Ash (%)	0202	6.185	0.0300	6.125	0.3238	0.1172	130	0.18	005.99	0
005	Ash (%)	2184	6.195	0.0500	6.125	0.3238	0.1172	130	0.22	005.00	0
005	Ash (%)	2246	6.200	0.0200	6.125	0.3238	0.1172	130	0.23	005.00	0
005	Ash (%)	0692	6.205	0.0700	6.125	0.3238	0.1172	130	0.25	005.05	0
005	Ash (%)	0529	6.220	0.3600	6.125	0.3238	0.1172	130	0.29	005.00	0
005	Ash (%)	0541	6.220	0.4400	6.125	0.3238	0.1172	130	0.29	005.00	0
005	Ash (%)	0939	6.235	0.1100	6.125	0.3238	0.1172	130	0.34	005.05	0
005	Ash (%)	0574	6.240	0.1200	6.125	0.3238	0.1172	130	0.35	005.00	0
005	Ash (%)	0860	6.240	0.0000	6.125	0.3238	0.1172	130	0.35	005.00	0
005	Ash (%)	0265	6.250	0.1000	6.125	0.3238	0.1172	130	0.38	005.00	0
005	Ash (%)	2188	6.260	0.0000	6.125	0.3238	0.1172	130	0.42	005.00	0
005	Ash (%)	0646	6.265	0.2500	6.125	0.3238	0.1172	130	0.43	005.00	0
005	Ash (%)	2009	6.287	0.0107	6.125	0.3238	0.1172	130	0.50	005.05	0
005	Ash (%)	2200	6.300	0.2000	6.125	0.3238	0.1172	130	0.54	005.00	0
005	Ash (%)	0948	6.320	0.0200	6.125	0.3238	0.1172	130	0.60	005.00	0
005	Ash (%)	2319	6.350	0.7000	6.125	0.3238	0.1172	130	0.69	005.99	0
005	Ash (%)	0083	6.360	0.0200	6.125	0.3238	0.1172	130	0.72	005.00	0
005	Ash (%)	2073	6.360	0.1000	6.125	0.3238	0.1172	130	0.72	005.00	0
005	Ash (%)	0186	6.361	0.1740	6.125	0.3238	0.1172	130	0.73	005.00	0
005	Ash (%)	2105	6.365	0.0300	6.125	0.3238	0.1172	130	0.74	005.11	0
005	Ash (%)	2181	6.380	0.3400	6.125	0.3238	0.1172	130	0.79	005.00	0
005	Ash (%)	2259	6.402	0.0790	6.125	0.3238	0.1172	130	0.85	005.00	0
005	Ash (%)	0682	6.420	0.0000	6.125	0.3238	0.1172	130	0.91	005.00	0
005	Ash (%)	0964	6.441	0.2840	6.125	0.3238	0.1172	130	0.97	005.00	0
005	Ash (%)	0300	6.455	0.1100	6.125	0.3238	0.1172	130	1.02	005.00	0
005	Ash (%)	2097	6.456	0.2943	6.125	0.3238	0.1172	130	1.02	005.00	0
005	Ash (%)	0941	6.465	0.1900	6.125	0.3238	0.1172	130	1.05	005.05	0
005	Ash (%)	0353	6.480	0.0200	6.125	0.3238	0.1172	130	1.10	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
005	Ash (%)	0653	6.490	0.0200	6.125	0.3238	0.1172	130	1.13	005.05	0
005	Ash (%)	2290	6.500	0.2000	6.125	0.3238	0.1172	130	1.16	005.11	0
005	Ash (%)	0190	6.545	0.0700	6.125	0.3238	0.1172	130	1.30	005.05	0
005	Ash (%)	2129	6.565	0.6700	6.125	0.3238	0.1172	130	1.36	005.05	0
005	Ash (%)	0100	6.600	0.0600	6.125	0.3238	0.1172	130	1.47	005.00	0
005	Ash (%)	0357	6.600	0.2000	6.125	0.3238	0.1172	130	1.47	005.00	0
005	Ash (%)	0366	6.600	0.2000	6.125	0.3238	0.1172	130	1.47	005.00	0
005	Ash (%)	0921	6.617	0.0760	6.125	0.3238	0.1172	130	1.52	005.05	0
005	Ash (%)	0297	6.630	0.0600	6.125	0.3238	0.1172	130	1.56	005.11	0
005	Ash (%)	0853	6.650	0.3000	6.125	0.3238	0.1172	130	1.62	005.00	0
005	Ash (%)	0536	6.655	0.0100	6.125	0.3238	0.1172	130	1.64	005.99	0
005	Ash (%)	0309	6.667	0.1380	6.125	0.3238	0.1172	130	1.67	005.00	0
005	Ash (%)	0123	6.685	0.2100	6.125	0.3238	0.1172	130	1.73	005.00	0
005	Ash (%)	2058	6.690	0.0600	6.125	0.3238	0.1172	130	1.74	005.00	0
005	Ash (%)	0164	6.700	0.1800	6.125	0.3238	0.1172	130	1.77	005.00	0
005	Ash (%)	0685	6.790	0.1800	6.125	0.3238	0.1172	130	2.05	005.05	0
005	Ash (%)	0003	6.795	0.0300	6.125	0.3238	0.1172	130	2.07	005.05	0
005	Ash (%)	2192	6.905	0.0900	6.125	0.3238	0.1172	130	2.41	005.05	0
005	Ash (%)	0852	7.000	0.0000	6.125	0.3238	0.1172	130	2.70	005.00	0
005	Ash (%)	0884	7.400	0.0000	6.125	0.3238	0.1172	130	3.94	005.99	0
005	Ash (%)	0510	8.285	0.0300	6.125	0.3238	0.1172	130	6.67	005.00	0
005	Ash (%)	0019	6.500	0.8000	6.125	0.3238	0.1172	130	1.16	005.00	1
005	Ash (%)	2076	6.862	1.344	6.125	0.3238	0.1172	130	2.27	005.00	1
005	Ash (%)	0108	7.320	1.060	6.125	0.3238	0.1172	130	3.69	005.00	1
006	Total Sugars (%)	0226	9.400	0.0600	11.09	1.570	0.3500	4		006.99	0
006	Total Sugars (%)	2290	10.60	0.4000	11.09	1.570	0.3500	4		006.99	0
006	Total Sugars (%)	0956	11.20	0.2000	11.09	1.570	0.3500	4		006.99	0
006	Total Sugars (%)	0407	13.16	0.7400	11.09	1.570	0.3500	4		006.00	0
008	Fiber, Acid Detergent (%)	0297	3.300	0.6000	4.927	0.9350	0.2366	53	-1.74	008.08	0
008	Fiber, Acid Detergent (%)	2345	3.320	0.2800	4.927	0.9350	0.2366	53	-1.72	008.08	0
008	Fiber, Acid Detergent (%)	0529	3.595	0.0700	4.927	0.9350	0.2366	53	-1.43	008.08	0
008	Fiber, Acid Detergent (%)	0674	3.600	0.0000	4.927	0.9350	0.2366	53	-1.42	008.02	0
008	Fiber, Acid Detergent (%)	0968	3.655	0.0300	4.927	0.9350	0.2366	53	-1.36	008.08	0
008	Fiber, Acid Detergent (%)	0956	3.700	0.2000	4.927	0.9350	0.2366	53	-1.31	008.08	0
008	Fiber, Acid Detergent (%)	0676	3.890	0.5000	4.927	0.9350	0.2366	53	-1.11	008.02	0
008	Fiber, Acid Detergent (%)	0653	4.005	0.0300	4.927	0.9350	0.2366	53	-0.99	008.08	0
008	Fiber, Acid Detergent (%)	2302	4.100	0.3200	4.927	0.9350	0.2366	53	-0.88	008.08	0
008	Fiber, Acid Detergent (%)	0726	4.200	0.0600	4.927	0.9350	0.2366	53	-0.78	008.08	0
008	Fiber, Acid Detergent (%)	0148	4.290	0.0200	4.927	0.9350	0.2366	53	-0.68	008.08	0
008	Fiber, Acid Detergent (%)	0354	4.360	0.0400	4.927	0.9350	0.2366	53	-0.61	008.08	0
008	Fiber, Acid Detergent (%)	0675	4.360	0.0400	4.927	0.9350	0.2366	53	-0.61	008.08	0
008	Fiber, Acid Detergent (%)	0345	4.450	0.3000	4.927	0.9350	0.2366	53	-0.51	008.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
008	Fiber, Acid Detergent (%)	0357	4.480	0.1800	4.927	0.9350	0.2366	53	-0.48	008.08	0
008	Fiber, Acid Detergent (%)	2192	4.485	0.3100	4.927	0.9350	0.2366	53	-0.47	008.02	0
008	Fiber, Acid Detergent (%)	0875	4.515	0.1300	4.927	0.9350	0.2366	53	-0.44	008.02	0
008	Fiber, Acid Detergent (%)	0407	4.586	0.2427	4.927	0.9350	0.2366	53	-0.37	008.08	0
008	Fiber, Acid Detergent (%)	0190	4.680	0.0800	4.927	0.9350	0.2366	53	-0.26	008.08	0
008	Fiber, Acid Detergent (%)	0309	4.694	0.0668	4.927	0.9350	0.2366	53	-0.25	008.02	0
008	Fiber, Acid Detergent (%)	0098	4.695	0.2300	4.927	0.9350	0.2366	53	-0.25	008.02	0
008	Fiber, Acid Detergent (%)	0511	4.700	0.4000	4.927	0.9350	0.2366	53	-0.24	008.08	0
008	Fiber, Acid Detergent (%)	0004	4.755	0.1300	4.927	0.9350	0.2366	53	-0.18	008.08	0
008	Fiber, Acid Detergent (%)	0148	4.775	0.0300	4.927	0.9350	0.2366	53	-0.16	008.02	0
008	Fiber, Acid Detergent (%)	0948	4.795	0.0700	4.927	0.9350	0.2366	53	-0.14	008.02	0
008	Fiber, Acid Detergent (%)	0910	4.830	0.4600	4.927	0.9350	0.2366	53	-0.10	008.08	0
008	Fiber, Acid Detergent (%)	0964	4.836	0.2310	4.927	0.9350	0.2366	53	-0.10	008.02	0
008	Fiber, Acid Detergent (%)	0035	4.840	0.3800	4.927	0.9350	0.2366	53	-0.09	008.08	0
008	Fiber, Acid Detergent (%)	0027	4.853	0.1730	4.927	0.9350	0.2366	53	-0.08	008.99	0
008	Fiber, Acid Detergent (%)	0164	4.900	0.0000	4.927	0.9350	0.2366	53	-0.03	008.08	0
008	Fiber, Acid Detergent (%)	2246	4.935	0.4100	4.927	0.9350	0.2366	53	0.01	008.02	0
008	Fiber, Acid Detergent (%)	0353	4.945	0.2100	4.927	0.9350	0.2366	53	0.02	008.02	0
008	Fiber, Acid Detergent (%)	0263	4.957	0.1730	4.927	0.9350	0.2366	53	0.03	008.08	0
008	Fiber, Acid Detergent (%)	0226	4.985	0.0900	4.927	0.9350	0.2366	53	0.06	008.02	0
008	Fiber, Acid Detergent (%)	0042	5.010	0.1000	4.927	0.9350	0.2366	53	0.09	008.08	0
008	Fiber, Acid Detergent (%)	2290	5.200	0.2000	4.927	0.9350	0.2366	53	0.29	008.99	0
008	Fiber, Acid Detergent (%)	0870	5.270	0.0207	4.927	0.9350	0.2366	53	0.37	008.08	0
008	Fiber, Acid Detergent (%)	0689	5.300	0.2000	4.927	0.9350	0.2366	53	0.40	008.02	0
008	Fiber, Acid Detergent (%)	0265	5.325	0.6100	4.927	0.9350	0.2366	53	0.43	008.05	0
008	Fiber, Acid Detergent (%)	0066	5.495	0.4700	4.927	0.9350	0.2366	53	0.61	008.08	0
008	Fiber, Acid Detergent (%)	2193	5.530	0.0400	4.927	0.9350	0.2366	53	0.64	008.08	0
008	Fiber, Acid Detergent (%)	0278	5.650	0.3000	4.927	0.9350	0.2366	53	0.77	008.08	0
008	Fiber, Acid Detergent (%)	0510	5.700	0.2000	4.927	0.9350	0.2366	53	0.83	008.08	0
008	Fiber, Acid Detergent (%)	0037	5.780	0.7400	4.927	0.9350	0.2366	53	0.91	008.08	0
008	Fiber, Acid Detergent (%)	0083	5.905	0.1100	4.927	0.9350	0.2366	53	1.05	008.08	0
008	Fiber, Acid Detergent (%)	2326	6.225	0.0500	4.927	0.9350	0.2366	53	1.39	008.99	0
008	Fiber, Acid Detergent (%)	0939	6.290	1.000	4.927	0.9350	0.2366	53	1.46	008.08	0
008	Fiber, Acid Detergent (%)	0045	6.640	0.6400	4.927	0.9350	0.2366	53	1.83	008.08	0
008	Fiber, Acid Detergent (%)	0581	6.850	0.1000	4.927	0.9350	0.2366	53	2.06	008.08	0
008	Fiber, Acid Detergent (%)	0693	6.971	0.1420	4.927	0.9350	0.2366	53	2.19	008.08	0
008	Fiber, Acid Detergent (%)	0049	7.350	0.1600	4.927	0.9350	0.2366	53	2.59	008.08	0
008	Fiber, Acid Detergent (%)	0504	7.655	0.1500	4.927	0.9350	0.2366	53	2.92	008.02	0
008	Fiber, Acid Detergent (%)	2129	9.910	0.8200	4.927	0.9350	0.2366	53	5.33	008.08	0
008	Fiber, Acid Detergent (%)	0848	5.395	1.390	4.927	0.9350	0.2366	53	0.50	008.08	1
009	Fiber, Neutral Detergent (%)	0638	4.900	0.0000	7.765	1.110	0.3731	50	-2.58	009.09	0
009	Fiber, Neutral Detergent (%)	0297	6.000	0.4000	7.765	1.110	0.3731	50	-1.59	009.09	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
009	Fiber, Neutral Detergent (%)	0345	6.400	0.4000	7.765	1.110	0.3731	50	-1.23	009.07	0
009	Fiber, Neutral Detergent (%)	0353	6.475	0.1100	7.765	1.110	0.3731	50	-1.16	009.07	0
009	Fiber, Neutral Detergent (%)	0956	6.700	0.2000	7.765	1.110	0.3731	50	-0.96	009.09	0
009	Fiber, Neutral Detergent (%)	0407	6.764	0.2223	7.765	1.110	0.3731	50	-0.90	009.09	0
009	Fiber, Neutral Detergent (%)	0263	6.808	0.0070	7.765	1.110	0.3731	50	-0.86	009.09	0
009	Fiber, Neutral Detergent (%)	0939	6.820	0.4600	7.765	1.110	0.3731	50	-0.85	009.09	0
009	Fiber, Neutral Detergent (%)	2345	6.820	0.4200	7.765	1.110	0.3731	50	-0.85	009.09	0
009	Fiber, Neutral Detergent (%)	0148	6.835	0.1900	7.765	1.110	0.3731	50	-0.84	009.09	0
009	Fiber, Neutral Detergent (%)	2022	6.840	0.1000	7.765	1.110	0.3731	50	-0.83	009.09	0
009	Fiber, Neutral Detergent (%)	0164	7.000	0.0000	7.765	1.110	0.3731	50	-0.69	009.09	0
009	Fiber, Neutral Detergent (%)	0726	7.000	0.4800	7.765	1.110	0.3731	50	-0.69	009.09	0
009	Fiber, Neutral Detergent (%)	0035	7.115	0.0100	7.765	1.110	0.3731	50	-0.59	009.09	0
009	Fiber, Neutral Detergent (%)	0870	7.176	0.1553	7.765	1.110	0.3731	50	-0.53	009.09	0
009	Fiber, Neutral Detergent (%)	0066	7.185	0.4500	7.765	1.110	0.3731	50	-0.52	009.09	0
009	Fiber, Neutral Detergent (%)	0278	7.200	0.2000	7.765	1.110	0.3731	50	-0.51	009.09	0
009	Fiber, Neutral Detergent (%)	0226	7.210	0.0800	7.765	1.110	0.3731	50	-0.50	009.07	0
009	Fiber, Neutral Detergent (%)	0098	7.270	0.2200	7.765	1.110	0.3731	50	-0.45	009.07	0
009	Fiber, Neutral Detergent (%)	0598	7.280	0.0000	7.765	1.110	0.3731	50	-0.44	009.09	0
009	Fiber, Neutral Detergent (%)	0581	7.300	0.2000	7.765	1.110	0.3731	50	-0.42	009.09	0
009	Fiber, Neutral Detergent (%)	0265	7.340	0.6800	7.765	1.110	0.3731	50	-0.38	009.09	0
009	Fiber, Neutral Detergent (%)	0948	7.345	0.0700	7.765	1.110	0.3731	50	-0.38	009.07	0
009	Fiber, Neutral Detergent (%)	2302	7.465	0.2300	7.765	1.110	0.3731	50	-0.27	009.09	0
009	Fiber, Neutral Detergent (%)	2192	7.535	0.2700	7.765	1.110	0.3731	50	-0.21	009.07	0
009	Fiber, Neutral Detergent (%)	0693	7.565	1.116	7.765	1.110	0.3731	50	-0.18	009.09	0
009	Fiber, Neutral Detergent (%)	0357	7.605	0.1300	7.765	1.110	0.3731	50	-0.14	009.09	0
009	Fiber, Neutral Detergent (%)	2246	7.620	0.5600	7.765	1.110	0.3731	50	-0.13	009.04	0
009	Fiber, Neutral Detergent (%)	0964	7.718	1.085	7.765	1.110	0.3731	50	-0.04	009.07	0
009	Fiber, Neutral Detergent (%)	0674	7.820	0.0000	7.765	1.110	0.3731	50	0.05	009.07	0
009	Fiber, Neutral Detergent (%)	0910	8.000	0.6000	7.765	1.110	0.3731	50	0.21	009.07	0
009	Fiber, Neutral Detergent (%)	0504	8.005	0.4300	7.765	1.110	0.3731	50	0.22	009.04	0
009	Fiber, Neutral Detergent (%)	0875	8.055	0.1700	7.765	1.110	0.3731	50	0.26	009.07	0
009	Fiber, Neutral Detergent (%)	0309	8.130	0.0595	7.765	1.110	0.3731	50	0.33	009.07	0
009	Fiber, Neutral Detergent (%)	0676	8.275	0.6900	7.765	1.110	0.3731	50	0.46	009.07	0
009	Fiber, Neutral Detergent (%)	2246	8.325	0.8500	7.765	1.110	0.3731	50	0.50	009.07	0
009	Fiber, Neutral Detergent (%)	0083	8.420	0.2400	7.765	1.110	0.3731	50	0.59	009.09	0
009	Fiber, Neutral Detergent (%)	2193	8.420	0.1800	7.765	1.110	0.3731	50	0.59	009.09	0
009	Fiber, Neutral Detergent (%)	0968	8.615	0.4700	7.765	1.110	0.3731	50	0.77	009.09	0
009	Fiber, Neutral Detergent (%)	0653	8.630	0.4200	7.765	1.110	0.3731	50	0.78	009.09	0
009	Fiber, Neutral Detergent (%)	2326	8.830	0.1000	7.765	1.110	0.3731	50	0.96	009.99	0
009	Fiber, Neutral Detergent (%)	0689	8.900	0.2000	7.765	1.110	0.3731	50	1.02	009.07	0
009	Fiber, Neutral Detergent (%)	0675	9.205	0.1900	7.765	1.110	0.3731	50	1.30	009.09	0
009	Fiber, Neutral Detergent (%)	0037	9.450	1.980	7.765	1.110	0.3731	50	1.52	009.09	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
009	Fiber, Neutral Detergent (%)	0510	9.450	0.1000	7.765	1.110	0.3731	50	1.52	009.09	0
009	Fiber, Neutral Detergent (%)	0354	9.590	0.3400	7.765	1.110	0.3731	50	1.64	009.09	0
009	Fiber, Neutral Detergent (%)	0049	10.11	0.9400	7.765	1.110	0.3731	50	2.11	009.09	0
009	Fiber, Neutral Detergent (%)	0190	10.19	0.3400	7.765	1.110	0.3731	50	2.18	009.09	0
009	Fiber, Neutral Detergent (%)	0045	10.35	1.310	7.765	1.110	0.3731	50	2.32	009.09	0
009	Fiber, Neutral Detergent (%)	2290	10.60	0.6000	7.765	1.110	0.3731	50	2.55	009.99	0
009	Fiber, Neutral Detergent (%)	0848	19.92	5.180	7.765	1.110	0.3731	50	10.95	009.09	1
009	Fiber, Neutral Detergent (%)	2129	26.70	5.200	7.765	1.110	0.3731	50	17.06	009.09	2
010	Moisture (%)	2105	9.630	0.0200	11.11	0.4666	0.1308	24	-3.18	010.11	0
010	Moisture (%)	0843	10.21	0.0100	11.11	0.4666	0.1308	24	-1.95	010.03	0
010	Moisture (%)	0011	10.40	0.0900	11.11	0.4666	0.1308	24	-1.54	010.99	0
010	Moisture (%)	0546	10.46	0.0700	11.11	0.4666	0.1308	24	-1.41	010.99	0
010	Moisture (%)	2129	10.64	0.5200	11.11	0.4666	0.1308	24	-1.02	010.99	0
010	Moisture (%)	2043	10.87	0.0600	11.11	0.4666	0.1308	24	-0.52	010.99	0
010	Moisture (%)	0852	10.93	0.1300	11.11	0.4666	0.1308	24	-0.40	010.11	0
010	Moisture (%)	0948	10.99	0.0100	11.11	0.4666	0.1308	24	-0.28	010.99	0
010	Moisture (%)	2076	11.09	0.2900	11.11	0.4666	0.1308	24	-0.06	010.99	0
010	Moisture (%)	0964	11.10	0.1280	11.11	0.4666	0.1308	24	-0.02	010.99	0
010	Moisture (%)	2089	11.11	0.0700	11.11	0.4666	0.1308	24	-0.02	010.99	0
010	Moisture (%)	0921	11.16	0.0400	11.11	0.4666	0.1308	24	0.10	010.99	0
010	Moisture (%)	0875	11.19	0.0200	11.11	0.4666	0.1308	24	0.16	010.99	0
010	Moisture (%)	0939	11.19	0.1800	11.11	0.4666	0.1308	24	0.16	010.99	0
010	Moisture (%)	2290	11.23	0.0600	11.11	0.4666	0.1308	24	0.25	010.11	0
010	Moisture (%)	2178	11.29	0.2900	11.11	0.4666	0.1308	24	0.37	010.99	0
010	Moisture (%)	0726	11.35	0.0300	11.11	0.4666	0.1308	24	0.50	010.99	0
010	Moisture (%)	0652	11.35	0.1000	11.11	0.4666	0.1308	24	0.51	010.99	0
010	Moisture (%)	2150	11.44	0.0100	11.11	0.4666	0.1308	24	0.69	010.99	0
010	Moisture (%)	0914	11.45	0.1800	11.11	0.4666	0.1308	24	0.72	010.99	0
010	Moisture (%)	0190	11.59	0.0000	11.11	0.4666	0.1308	24	1.02	010.99	0
010	Moisture (%)	0164	11.63	0.1200	11.11	0.4666	0.1308	24	1.11	010.03	0
010	Moisture (%)	0884	11.65	0.7000	11.11	0.4666	0.1308	24	1.15	010.99	0
010	Moisture (%)	2329	11.99	0.0100	11.11	0.4666	0.1308	24	1.87	010.99	0
010	Moisture (%)	0692	8.045	5.930	11.11	0.4666	0.1308	24	-6.58	010.99	1
011	Loss on Drying, HT (%)	0563	10.12	0.0470	11.82	0.2457	0.1064	71	-6.89	011.01	0
011	Loss on Drying, HT (%)	2199	10.60	0.2000	11.82	0.2457	0.1064	71	-4.95	011.01	0
011	Loss on Drying, HT (%)	0651	10.70	0.1703	11.82	0.2457	0.1064	71	-4.55	011.01	0
011	Loss on Drying, HT (%)	0859	10.98	0.0980	11.82	0.2457	0.1064	71	-3.41	011.01	0
011	Loss on Drying, HT (%)	2200	11.05	0.1000	11.82	0.2457	0.1064	71	-3.11	011.01	0
011	Loss on Drying, HT (%)	0953	11.29	0.0300	11.82	0.2457	0.1064	71	-2.16	011.01	0
011	Loss on Drying, HT (%)	2188	11.33	0.0600	11.82	0.2457	0.1064	71	-1.97	011.01	0
011	Loss on Drying, HT (%)	0108	11.49	0.3700	11.82	0.2457	0.1064	71	-1.34	011.01	0
011	Loss on Drying, HT (%)	0852	11.55	0.1000	11.82	0.2457	0.1064	71	-1.08	011.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
011	Loss on Drying, HT (%)	0598	11.56	0.0300	11.82	0.2457	0.1064	71	-1.06	011.01	0
011	Loss on Drying, HT (%)	0226	11.59	0.0500	11.82	0.2457	0.1064	71	-0.94	011.01	0
011	Loss on Drying, HT (%)	2193	11.59	0.0500	11.82	0.2457	0.1064	71	-0.94	011.01	0
011	Loss on Drying, HT (%)	0265	11.60	0.0000	11.82	0.2457	0.1064	71	-0.88	011.01	0
011	Loss on Drying, HT (%)	0354	11.66	0.0100	11.82	0.2457	0.1064	71	-0.65	011.01	0
011	Loss on Drying, HT (%)	0511	11.66	0.1100	11.82	0.2457	0.1064	71	-0.65	011.01	0
011	Loss on Drying, HT (%)	0066	11.67	0.1300	11.82	0.2457	0.1064	71	-0.61	011.01	0
011	Loss on Drying, HT (%)	0685	11.67	0.4000	11.82	0.2457	0.1064	71	-0.59	011.01	0
011	Loss on Drying, HT (%)	0263	11.68	0.0320	11.82	0.2457	0.1064	71	-0.56	011.01	0
011	Loss on Drying, HT (%)	0553	11.68	0.1400	11.82	0.2457	0.1064	71	-0.55	011.01	0
011	Loss on Drying, HT (%)	0968	11.70	0.0100	11.82	0.2457	0.1064	71	-0.49	011.01	0
011	Loss on Drying, HT (%)	0208	11.70	0.0000	11.82	0.2457	0.1064	71	-0.47	011.01	0
011	Loss on Drying, HT (%)	0510	11.70	0.4000	11.82	0.2457	0.1064	71	-0.47	011.01	0
011	Loss on Drying, HT (%)	0123	11.71	0.1400	11.82	0.2457	0.1064	71	-0.43	011.01	0
011	Loss on Drying, HT (%)	2181	11.71	0.1400	11.82	0.2457	0.1064	71	-0.43	011.01	0
011	Loss on Drying, HT (%)	2268	11.72	0.3700	11.82	0.2457	0.1064	71	-0.41	011.01	0
011	Loss on Drying, HT (%)	0309	11.73	0.3100	11.82	0.2457	0.1064	71	-0.37	011.01	0
011	Loss on Drying, HT (%)	0229	11.73	0.3400	11.82	0.2457	0.1064	71	-0.35	011.01	0
011	Loss on Drying, HT (%)	0536	11.74	0.0100	11.82	0.2457	0.1064	71	-0.33	011.01	0
011	Loss on Drying, HT (%)	0574	11.74	0.0100	11.82	0.2457	0.1064	71	-0.33	011.01	0
011	Loss on Drying, HT (%)	0646	11.74	0.0100	11.82	0.2457	0.1064	71	-0.33	011.01	0
011	Loss on Drying, HT (%)	2174	11.74	0.0900	11.82	0.2457	0.1064	71	-0.33	011.01	0
011	Loss on Drying, HT (%)	0144	11.76	0.1200	11.82	0.2457	0.1064	71	-0.22	011.01	0
011	Loss on Drying, HT (%)	2303	11.77	0.0160	11.82	0.2457	0.1064	71	-0.19	011.99	0
011	Loss on Drying, HT (%)	0942	11.78	0.0700	11.82	0.2457	0.1064	71	-0.16	011.02	0
011	Loss on Drying, HT (%)	2073	11.79	0.2500	11.82	0.2457	0.1064	71	-0.12	011.01	0
011	Loss on Drying, HT (%)	2174	11.80	0.2000	11.82	0.2457	0.1064	71	-0.06	011.99	0
011	Loss on Drying, HT (%)	0653	11.81	0.0500	11.82	0.2457	0.1064	71	-0.04	011.01	0
011	Loss on Drying, HT (%)	0100	11.82	0.0200	11.82	0.2457	0.1064	71	0.02	011.01	0
011	Loss on Drying, HT (%)	0520	11.84	0.0200	11.82	0.2457	0.1064	71	0.10	011.01	0
011	Loss on Drying, HT (%)	0529	11.84	0.0000	11.82	0.2457	0.1064	71	0.10	011.02	0
011	Loss on Drying, HT (%)	0171	11.85	0.0300	11.82	0.2457	0.1064	71	0.12	011.01	0
011	Loss on Drying, HT (%)	0425	11.87	0.0300	11.82	0.2457	0.1064	71	0.20	011.01	0
011	Loss on Drying, HT (%)	0242	11.88	0.0800	11.82	0.2457	0.1064	71	0.26	011.01	0
011	Loss on Drying, HT (%)	0098	11.89	0.0700	11.82	0.2457	0.1064	71	0.28	011.01	0
011	Loss on Drying, HT (%)	0541	11.89	0.1900	11.82	0.2457	0.1064	71	0.28	011.01	0
011	Loss on Drying, HT (%)	0674	11.90	0.0000	11.82	0.2457	0.1064	71	0.35	011.01	0
011	Loss on Drying, HT (%)	2168	11.93	0.2900	11.82	0.2457	0.1064	71	0.45	011.01	0
011	Loss on Drying, HT (%)	2326	11.93	0.0300	11.82	0.2457	0.1064	71	0.45	011.01	0
011	Loss on Drying, HT (%)	0164	11.94	0.1700	11.82	0.2457	0.1064	71	0.49	011.01	0
011	Loss on Drying, HT (%)	0643	11.95	0.0500	11.82	0.2457	0.1064	71	0.53	011.01	0
011	Loss on Drying, HT (%)	0300	11.95	0.3000	11.82	0.2457	0.1064	71	0.55	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
011	Loss on Drying, HT (%)	2097	11.96	0.0407	11.82	0.2457	0.1064	71	0.58	011.01	0
011	Loss on Drying, HT (%)	0539	11.96	0.0600	11.82	0.2457	0.1064	71	0.59	011.01	0
011	Loss on Drying, HT (%)	0589	11.96	0.0400	11.82	0.2457	0.1064	71	0.59	011.01	0
011	Loss on Drying, HT (%)	2006	11.96	0.1400	11.82	0.2457	0.1064	71	0.59	011.01	0
011	Loss on Drying, HT (%)	2302	11.99	0.0900	11.82	0.2457	0.1064	71	0.69	011.01	0
011	Loss on Drying, HT (%)	2168	12.00	0.1400	11.82	0.2457	0.1064	71	0.75	011.99	0
011	Loss on Drying, HT (%)	0682	12.01	0.0000	11.82	0.2457	0.1064	71	0.79	011.01	0
011	Loss on Drying, HT (%)	0004	12.02	0.0900	11.82	0.2457	0.1064	71	0.81	011.01	0
011	Loss on Drying, HT (%)	0870	12.02	0.1555	11.82	0.2457	0.1064	71	0.83	011.01	0
011	Loss on Drying, HT (%)	0407	12.10	0.0125	11.82	0.2457	0.1064	71	1.17	011.01	0
011	Loss on Drying, HT (%)	0618	12.11	0.0200	11.82	0.2457	0.1064	71	1.20	011.01	0
011	Loss on Drying, HT (%)	2058	12.11	0.0000	11.82	0.2457	0.1064	71	1.20	011.01	0
011	Loss on Drying, HT (%)	0202	12.15	0.1000	11.82	0.2457	0.1064	71	1.36	011.01	0
011	Loss on Drying, HT (%)	2184	12.26	0.0800	11.82	0.2457	0.1064	71	1.81	011.01	0
011	Loss on Drying, HT (%)	2246	12.27	0.0400	11.82	0.2457	0.1064	71	1.85	011.01	0
011	Loss on Drying, HT (%)	0675	12.31	0.0200	11.82	0.2457	0.1064	71	2.01	011.01	0
011	Loss on Drying, HT (%)	0047	12.55	0.3000	11.82	0.2457	0.1064	71	2.99	011.01	0
011	Loss on Drying, HT (%)	2337	12.59	0.0100	11.82	0.2457	0.1064	71	3.13	011.01	0
011	Loss on Drying, HT (%)	0843	12.65	0.1100	11.82	0.2457	0.1064	71	3.38	011.01	0
011	Loss on Drying, HT (%)	0062	12.98	0.1600	11.82	0.2457	0.1064	71	4.74	011.01	0
011	Loss on Drying, HT (%)	2055	10.50	1.000	11.82	0.2457	0.1064	71	-5.35	011.99	1
012	Starch (%)	2290	0.6500	0.1000	1.596	0.7398	0.1110	18	-1.28	012.11	0
012	Starch (%)	0407	0.8678	0.0045	1.596	0.7398	0.1110	18	-0.98	012.01	0
012	Starch (%)	0164	0.9500	0.1000	1.596	0.7398	0.1110	18	-0.87	012.01	0
012	Starch (%)	0407	1.049	0.0560	1.596	0.7398	0.1110	18	-0.74	012.03	0
012	Starch (%)	0226	1.125	0.0100	1.596	0.7398	0.1110	18	-0.64	012.04	0
012	Starch (%)	0956	1.150	0.1000	1.596	0.7398	0.1110	18	-0.60	012.01	0
012	Starch (%)	0148	1.200	0.1400	1.596	0.7398	0.1110	18	-0.54	012.01	0
012	Starch (%)	0045	1.400	0.2000	1.596	0.7398	0.1110	18	-0.27	012.01	0
012	Starch (%)	0870	1.408	0.0089	1.596	0.7398	0.1110	18	-0.25	012.01	0
012	Starch (%)	0297	1.530	0.0200	1.596	0.7398	0.1110	18	-0.09	012.11	0
012	Starch (%)	0510	1.600	0.2000	1.596	0.7398	0.1110	18	0.01	012.04	0
012	Starch (%)	0278	1.700	0.2000	1.596	0.7398	0.1110	18	0.14	012.04	0
012	Starch (%)	0297	1.775	0.2300	1.596	0.7398	0.1110	18	0.24	012.03	0
012	Starch (%)	0693	1.795	0.0490	1.596	0.7398	0.1110	18	0.27	012.01	0
012	Starch (%)	2129	2.430	0.2400	1.596	0.7398	0.1110	18	1.13	012.04	0
012	Starch (%)	2089	2.690	0.2000	1.596	0.7398	0.1110	18	1.48	012.01	0
012	Starch (%)	0948	5.865	0.0300	1.596	0.7398	0.1110	18	5.77	012.04	0
012	Starch (%)	0353	6.475	0.1100	1.596	0.7398	0.1110	18	6.59	012.20	0
013	Fat, Acid Pretreat (%)	0948	0.2850	0.0500	2.312	0.9593	0.1513	47	-2.11	013.00	0
013	Fat, Acid Pretreat (%)	0638	0.6000	0.0000	2.312	0.9593	0.1513	47	-1.78	013.10	0
013	Fat, Acid Pretreat (%)	0036	0.7514	0.1484	2.312	0.9593	0.1513	47	-1.63	013.10	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
013	Fat, Acid Pretreat (%)	2268	0.9200	0.0200	2.312	0.9593	0.1513	47	-1.45	013.10	0
013	Fat, Acid Pretreat (%)	0353	0.9950	0.0900	2.312	0.9593	0.1513	47	-1.37	013.10	0
013	Fat, Acid Pretreat (%)	0683	1.145	0.0900	2.312	0.9593	0.1513	47	-1.22	013.13	0
013	Fat, Acid Pretreat (%)	0910	1.240	0.3200	2.312	0.9593	0.1513	47	-1.12	013.00	0
013	Fat, Acid Pretreat (%)	0968	1.300	0.2200	2.312	0.9593	0.1513	47	-1.05	013.13	0
013	Fat, Acid Pretreat (%)	0689	1.300	0.0000	2.312	0.9593	0.1513	47	-1.05	013.00	0
013	Fat, Acid Pretreat (%)	0618	1.562	0.1862	2.312	0.9593	0.1513	47	-0.78	013.08	0
013	Fat, Acid Pretreat (%)	0675	1.590	0.2800	2.312	0.9593	0.1513	47	-0.75	013.02	0
013	Fat, Acid Pretreat (%)	2345	1.660	0.0800	2.312	0.9593	0.1513	47	-0.68	013.00	0
013	Fat, Acid Pretreat (%)	2319	1.750	0.1000	2.312	0.9593	0.1513	47	-0.59	013.00	0
013	Fat, Acid Pretreat (%)	0581	1.800	0.2000	2.312	0.9593	0.1513	47	-0.53	013.13	0
013	Fat, Acid Pretreat (%)	2073	1.890	0.1000	2.312	0.9593	0.1513	47	-0.44	013.00	0
013	Fat, Acid Pretreat (%)	2290	1.950	0.1000	2.312	0.9593	0.1513	47	-0.38	013.12	0
013	Fat, Acid Pretreat (%)	0682	1.970	0.0000	2.312	0.9593	0.1513	47	-0.36	013.02	0
013	Fat, Acid Pretreat (%)	0083	1.980	0.0200	2.312	0.9593	0.1513	47	-0.35	013.00	0
013	Fat, Acid Pretreat (%)	0242	2.020	0.1400	2.312	0.9593	0.1513	47	-0.30	013.00	0
013	Fat, Acid Pretreat (%)	0297	2.070	0.4000	2.312	0.9593	0.1513	47	-0.25	013.00	0
013	Fat, Acid Pretreat (%)	0650	2.265	0.3300	2.312	0.9593	0.1513	47	-0.05	013.00	0
013	Fat, Acid Pretreat (%)	0843	2.285	0.0900	2.312	0.9593	0.1513	47	-0.03	013.13	0
013	Fat, Acid Pretreat (%)	0914	2.360	0.1600	2.312	0.9593	0.1513	47	0.05	013.02	0
013	Fat, Acid Pretreat (%)	0098	2.380	0.0600	2.312	0.9593	0.1513	47	0.07	013.02	0
013	Fat, Acid Pretreat (%)	0643	2.390	0.3000	2.312	0.9593	0.1513	47	0.08	013.02	0
013	Fat, Acid Pretreat (%)	0652	2.400	0.0000	2.312	0.9593	0.1513	47	0.09	013.00	0
013	Fat, Acid Pretreat (%)	0148	2.400	0.0200	2.312	0.9593	0.1513	47	0.09	013.02	0
013	Fat, Acid Pretreat (%)	0853	2.500	0.2000	2.312	0.9593	0.1513	47	0.20	013.02	0
013	Fat, Acid Pretreat (%)	0354	2.655	0.5100	2.312	0.9593	0.1513	47	0.36	013.02	0
013	Fat, Acid Pretreat (%)	0870	2.760	0.0375	2.312	0.9593	0.1513	47	0.47	013.02	0
013	Fat, Acid Pretreat (%)	2058	2.760	0.2000	2.312	0.9593	0.1513	47	0.47	013.02	0
013	Fat, Acid Pretreat (%)	0171	2.825	0.4500	2.312	0.9593	0.1513	47	0.54	013.02	0
013	Fat, Acid Pretreat (%)	0265	2.825	0.2500	2.312	0.9593	0.1513	47	0.54	013.13	0
013	Fat, Acid Pretreat (%)	2076	2.895	0.0703	2.312	0.9593	0.1513	47	0.61	013.00	0
013	Fat, Acid Pretreat (%)	0229	2.915	0.2300	2.312	0.9593	0.1513	47	0.63	013.00	0
013	Fat, Acid Pretreat (%)	0164	2.970	0.0000	2.312	0.9593	0.1513	47	0.69	013.02	0
013	Fat, Acid Pretreat (%)	0676	3.030	0.1200	2.312	0.9593	0.1513	47	0.75	013.02	0
013	Fat, Acid Pretreat (%)	0541	3.075	0.0300	2.312	0.9593	0.1513	47	0.80	013.00	0
013	Fat, Acid Pretreat (%)	0202	3.135	0.2500	2.312	0.9593	0.1513	47	0.86	013.00	0
013	Fat, Acid Pretreat (%)	0553	3.265	0.0700	2.312	0.9593	0.1513	47	0.99	013.02	0
013	Fat, Acid Pretreat (%)	2259	3.352	0.1010	2.312	0.9593	0.1513	47	1.08	013.02	0
013	Fat, Acid Pretreat (%)	0407	3.381	0.0525	2.312	0.9593	0.1513	47	1.12	013.13	0
013	Fat, Acid Pretreat (%)	2150	3.390	0.0200	2.312	0.9593	0.1513	47	1.12	013.00	0
013	Fat, Acid Pretreat (%)	0309	3.539	0.1103	2.312	0.9593	0.1513	47	1.28	013.00	0
013	Fat, Acid Pretreat (%)	0618	3.632	0.3205	2.312	0.9593	0.1513	47	1.38	013.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
013	Fat, Acid Pretreat (%)	0407	3.795	0.5700	2.312	0.9593	0.1513	47	1.55	013.02	0
013	Fat, Acid Pretreat (%)	0921	4.067	0.0160	2.312	0.9593	0.1513	47	1.83	013.02	0
015	Aluminum (ppm)	0037	5.150	5.100	15.92	3.810	1.750	15	-2.83	015.42	0
015	Aluminum (ppm)	0297	6.000	2.000	15.92	3.810	1.750	15	-2.60	015.43	0
015	Aluminum (ppm)	0049	12.92	1.220	15.92	3.810	1.750	15	-0.79	015.41	0
015	Aluminum (ppm)	0164	14.00	0.0000	15.92	3.810	1.750	15	-0.50	015.41	0
015	Aluminum (ppm)	0407	14.13	0.0290	15.92	3.810	1.750	15	-0.47	015.53	0
015	Aluminum (ppm)	0553	15.50	1.000	15.92	3.810	1.750	15	-0.11	015.53	0
015	Aluminum (ppm)	0407	15.57	0.4465	15.92	3.810	1.750	15	-0.09	015.41	0
015	Aluminum (ppm)	0123	16.20	0.2000	15.92	3.810	1.750	15	0.07	015.41	0
015	Aluminum (ppm)	0345	16.60	0.4000	15.92	3.810	1.750	15	0.18	015.43	0
015	Aluminum (ppm)	0964	17.10	0.2000	15.92	3.810	1.750	15	0.31	015.43	0
015	Aluminum (ppm)	0169	17.30	0.8000	15.92	3.810	1.750	15	0.36	015.43	0
015	Aluminum (ppm)	0520	17.44	1.620	15.92	3.810	1.750	15	0.40	015.41	0
015	Aluminum (ppm)	0407	18.35	0.0369	15.92	3.810	1.750	15	0.64	015.43	0
015	Aluminum (ppm)	0353	59.50	9.000	15.92	3.810	1.750	15	11.44	015.43	0
015	Aluminum (ppm)	2129	238.6	4.200	15.92	3.810	1.750	15	58.45	015.42	0
015	Aluminum (ppm)	0171	64.76	26.53	15.92	3.810	1.750	15	12.82	015.41	1
015	Aluminum (ppm)	0425	< 10		15.92	3.810	1.750	15		015.43	5
017	Boron (ppm)	2129	6.200	0.2000	33.70	3.792	0.8381	17	-7.25	017.42	0
017	Boron (ppm)	0123	9.180	0.1000	33.70	3.792	0.8381	17	-6.47	017.44	0
017	Boron (ppm)	0297	21.00	2.000	33.70	3.792	0.8381	17	-3.35	017.43	0
017	Boron (ppm)	0226	30.76	0.3600	33.70	3.792	0.8381	17	-0.78	017.41	0
017	Boron (ppm)	0229	31.26	2.500	33.70	3.792	0.8381	17	-0.64	017.41	0
017	Boron (ppm)	0693	32.83	0.2880	33.70	3.792	0.8381	17	-0.23	017.42	0
017	Boron (ppm)	0345	33.55	0.1000	33.70	3.792	0.8381	17	-0.04	017.43	0
017	Boron (ppm)	0407	34.35	0.5257	33.70	3.792	0.8381	17	0.17	017.43	0
017	Boron (ppm)	0918	35.09	0.7100	33.70	3.792	0.8381	17	0.36	017.43	0
017	Boron (ppm)	0265	35.25	0.9000	33.70	3.792	0.8381	17	0.41	017.42	0
017	Boron (ppm)	0037	35.35	0.1000	33.70	3.792	0.8381	17	0.43	017.42	0
017	Boron (ppm)	0407	35.67	0.0458	33.70	3.792	0.8381	17	0.52	017.41	0
017	Boron (ppm)	0049	36.17	1.970	33.70	3.792	0.8381	17	0.65	017.41	0
017	Boron (ppm)	0045	36.35	2.500	33.70	3.792	0.8381	17	0.70	017.42	0
017	Boron (ppm)	0510	37.00	0.0000	33.70	3.792	0.8381	17	0.87	017.43	0
017	Boron (ppm)	0407	37.56	1.848	33.70	3.792	0.8381	17	1.02	017.53	0
017	Boron (ppm)	0083	37.75	0.1000	33.70	3.792	0.8381	17	1.07	017.43	0
017	Boron (ppm)	0353	48.50	5.000	33.70	3.792	0.8381	17	3.90	017.43	1
019	Calcium (%)	0921	0.0465	0.0010	0.3277	0.0222	0.0126	85	-12.69	019.31	0
019	Calcium (%)	2006	0.1905	0.0110	0.3277	0.0222	0.0126	85	-6.19	019.00	0
019	Calcium (%)	2006	0.2100	0.0200	0.3277	0.0222	0.0126	85	-5.31	019.09	0
019	Calcium (%)	0229	0.2800	0.0400	0.3277	0.0222	0.0126	85	-2.15	019.41	0
019	Calcium (%)	0968	0.2845	0.0030	0.3277	0.0222	0.0126	85	-1.95	019.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
019	Calcium (%)	0242	0.2900	0.0000	0.3277	0.0222	0.0126	85	-1.70	019.99	0
019	Calcium (%)	0407	0.2987	0.0025	0.3277	0.0222	0.0126	85	-1.31	019.53	0
019	Calcium (%)	0884	0.3000	0.0000	0.3277	0.0222	0.0126	85	-1.25	019.31	0
019	Calcium (%)	0171	0.3000	0.0000	0.3277	0.0222	0.0126	85	-1.25	019.41	0
019	Calcium (%)	0910	0.3000	0.0000	0.3277	0.0222	0.0126	85	-1.25	019.41	0
019	Calcium (%)	0425	0.3000	0.0400	0.3277	0.0222	0.0126	85	-1.25	019.43	0
019	Calcium (%)	0300	0.3014	0.0039	0.3277	0.0222	0.0126	85	-1.19	019.43	0
019	Calcium (%)	0870	0.3039	0.0024	0.3277	0.0222	0.0126	85	-1.07	019.42	0
019	Calcium (%)	0676	0.3050	0.0100	0.3277	0.0222	0.0126	85	-1.02	019.99	0
019	Calcium (%)	0870	0.3062	0.0028	0.3277	0.0222	0.0126	85	-0.97	019.43	0
019	Calcium (%)	2192	0.3100	0.0400	0.3277	0.0222	0.0126	85	-0.80	019.43	0
019	Calcium (%)	0033	0.3120	0.0020	0.3277	0.0222	0.0126	85	-0.71	019.43	0
019	Calcium (%)	2246	0.3150	0.0100	0.3277	0.0222	0.0126	85	-0.57	019.31	0
019	Calcium (%)	0004	0.3150	0.0100	0.3277	0.0222	0.0126	85	-0.57	019.41	0
019	Calcium (%)	0510	0.3150	0.0100	0.3277	0.0222	0.0126	85	-0.57	019.43	0
019	Calcium (%)	0511	0.3150	0.0100	0.3277	0.0222	0.0126	85	-0.57	019.43	0
019	Calcium (%)	0553	0.3150	0.0020	0.3277	0.0222	0.0126	85	-0.57	019.53	0
019	Calcium (%)	0598	0.3153	0.0024	0.3277	0.0222	0.0126	85	-0.56	019.41	0
019	Calcium (%)	0520	0.3165	0.0090	0.3277	0.0222	0.0126	85	-0.51	019.41	0
019	Calcium (%)	0876	0.3170	0.0120	0.3277	0.0222	0.0126	85	-0.48	019.43	0
019	Calcium (%)	0693	0.3185	0.0050	0.3277	0.0222	0.0126	85	-0.42	019.42	0
019	Calcium (%)	0186	0.3187	0.0118	0.3277	0.0222	0.0126	85	-0.41	019.52	0
019	Calcium (%)	0918	0.3190	0.0020	0.3277	0.0222	0.0126	85	-0.39	019.43	0
019	Calcium (%)	0045	0.3195	0.0310	0.3277	0.0222	0.0126	85	-0.37	019.42	0
019	Calcium (%)	0049	0.3200	0.0000	0.3277	0.0222	0.0126	85	-0.35	019.41	0
019	Calcium (%)	0098	0.3200	0.0000	0.3277	0.0222	0.0126	85	-0.35	019.41	0
019	Calcium (%)	0083	0.3200	0.0000	0.3277	0.0222	0.0126	85	-0.35	019.43	0
019	Calcium (%)	0297	0.3200	0.0200	0.3277	0.0222	0.0126	85	-0.35	019.43	0
019	Calcium (%)	0164	0.3200	0.0080	0.3277	0.0222	0.0126	85	-0.35	019.44	0
019	Calcium (%)	0563	0.3215	0.0130	0.3277	0.0222	0.0126	85	-0.28	019.31	0
019	Calcium (%)	0208	0.3215	0.0090	0.3277	0.0222	0.0126	85	-0.28	019.41	0
019	Calcium (%)	0653	0.3225	0.0010	0.3277	0.0222	0.0126	85	-0.23	019.41	0
019	Calcium (%)	0154	0.3243	0.0098	0.3277	0.0222	0.0126	85	-0.15	019.52	0
019	Calcium (%)	0263	0.3245	0.0070	0.3277	0.0222	0.0126	85	-0.14	019.42	0
019	Calcium (%)	0536	0.3250	0.0300	0.3277	0.0222	0.0126	85	-0.12	019.02	0
019	Calcium (%)	0529	0.3250	0.0100	0.3277	0.0222	0.0126	85	-0.12	019.31	0
019	Calcium (%)	0536	0.3250	0.0100	0.3277	0.0222	0.0126	85	-0.12	019.31	0
019	Calcium (%)	0144	0.3250	0.0100	0.3277	0.0222	0.0126	85	-0.12	019.41	0
019	Calcium (%)	0651	0.3260	0.0124	0.3277	0.0222	0.0126	85	-0.08	019.31	0
019	Calcium (%)	0572	0.3260	0.0720	0.3277	0.0222	0.0126	85	-0.08	019.53	0
019	Calcium (%)	0148	0.3260	0.0300	0.3277	0.0222	0.0126	85	-0.08	019.41	0
019	Calcium (%)	0345	0.3265	0.0070	0.3277	0.0222	0.0126	85	-0.05	019.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
019	Calcium (%)	0037	0.3285	0.0010	0.3277	0.0222	0.0126	85	0.04	019.42	0
019	Calcium (%)	2022	0.3300	0.0200	0.3277	0.0222	0.0126	85	0.10	019.00	0
019	Calcium (%)	0035	0.3300	0.0000	0.3277	0.0222	0.0126	85	0.10	019.42	0
019	Calcium (%)	0278	0.3300	0.0000	0.3277	0.0222	0.0126	85	0.10	019.42	0
019	Calcium (%)	0098	0.3300	0.0000	0.3277	0.0222	0.0126	85	0.10	019.43	0
019	Calcium (%)	0939	0.3300	0.0000	0.3277	0.0222	0.0126	85	0.10	019.53	0
019	Calcium (%)	0226	0.3350	0.0100	0.3277	0.0222	0.0126	85	0.33	019.41	0
019	Calcium (%)	0366	0.3350	0.0100	0.3277	0.0222	0.0126	85	0.33	019.42	0
019	Calcium (%)	0010	0.3350	0.0100	0.3277	0.0222	0.0126	85	0.33	019.43	0
019	Calcium (%)	0100	0.3350	0.0100	0.3277	0.0222	0.0126	85	0.33	019.99	0
019	Calcium (%)	0407	0.3370	0.0200	0.3277	0.0222	0.0126	85	0.42	019.41	0
019	Calcium (%)	0407	0.3384	0.0094	0.3277	0.0222	0.0126	85	0.48	019.43	0
019	Calcium (%)	0357	0.3389	0.0039	0.3277	0.0222	0.0126	85	0.50	019.42	0
019	Calcium (%)	2302	0.3400	0.0000	0.3277	0.0222	0.0126	85	0.56	019.00	0
019	Calcium (%)	0726	0.3400	0.0000	0.3277	0.0222	0.0126	85	0.56	019.42	0
019	Calcium (%)	0035	0.3400	0.0000	0.3277	0.0222	0.0126	85	0.56	019.43	0
019	Calcium (%)	2089	0.3400	0.0000	0.3277	0.0222	0.0126	85	0.56	019.43	0
019	Calcium (%)	0186	0.3435	0.0390	0.3277	0.0222	0.0126	85	0.71	019.42	0
019	Calcium (%)	0964	0.3447	0.0006	0.3277	0.0222	0.0126	85	0.77	019.43	0
019	Calcium (%)	0354	0.3450	0.0100	0.3277	0.0222	0.0126	85	0.78	019.41	0
019	Calcium (%)	0692	0.3450	0.0100	0.3277	0.0222	0.0126	85	0.78	019.42	0
019	Calcium (%)	0042	0.3490	0.0060	0.3277	0.0222	0.0126	85	0.96	019.42	0
019	Calcium (%)	0123	0.3500	0.0800	0.3277	0.0222	0.0126	85	1.01	019.41	0
019	Calcium (%)	0353	0.3500	0.0200	0.3277	0.0222	0.0126	85	1.01	019.43	0
019	Calcium (%)	2097	0.3539	0.0479	0.3277	0.0222	0.0126	85	1.18	019.31	0
019	Calcium (%)	0190	0.3550	0.0100	0.3277	0.0222	0.0126	85	1.23	019.42	0
019	Calcium (%)	0265	0.3550	0.0300	0.3277	0.0222	0.0126	85	1.23	019.42	0
019	Calcium (%)	0202	0.3550	0.0100	0.3277	0.0222	0.0126	85	1.23	019.43	0
019	Calcium (%)	2326	0.3550	0.0100	0.3277	0.0222	0.0126	85	1.23	019.43	0
019	Calcium (%)	2190	0.3555	0.0102	0.3277	0.0222	0.0126	85	1.25	019.08	0
019	Calcium (%)	2259	0.3565	0.0030	0.3277	0.0222	0.0126	85	1.30	019.41	0
019	Calcium (%)	0142	0.3637	0.0532	0.3277	0.0222	0.0126	85	1.62	019.31	0
019	Calcium (%)	0682	0.3700	0.0000	0.3277	0.0222	0.0126	85	1.91	019.43	0
019	Calcium (%)	2188	0.3800	0.0000	0.3277	0.0222	0.0126	85	2.36	019.08	0
019	Calcium (%)	0169	0.3800	0.0200	0.3277	0.0222	0.0126	85	2.36	019.32	0
019	Calcium (%)	0685	0.3950	0.0100	0.3277	0.0222	0.0126	85	3.04	019.41	0
019	Calcium (%)	2181	0.4250	0.0500	0.3277	0.0222	0.0126	85	4.39	019.00	0
019	Calcium (%)	2193	0.4350	0.0100	0.3277	0.0222	0.0126	85	4.84	019.08	0
019	Calcium (%)	2319	0.2255	0.1830	0.3277	0.0222	0.0126	85	-4.61	019.42	1
019	Calcium (%)	0852	0.2500	0.1000	0.3277	0.0222	0.0126	85	-3.51	019.08	1
019	Calcium (%)	0687	0.3000	0.1000	0.3277	0.0222	0.0126	85	-1.25	019.31	1
019	Calcium (%)	2129	5.500	0.0000	0.3277	0.0222	0.0126	85	233.47	019.42	2

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
021	Cobalt (ppm)	2326	0.1550	0.0300	0.3105	0.1267	0.0191	11	-1.23	021.43	0
021	Cobalt (ppm)	0123	0.2250	0.0500	0.3105	0.1267	0.0191	11	-0.67	021.41	0
021	Cobalt (ppm)	0171	0.2300	0.0020	0.3105	0.1267	0.0191	11	-0.64	021.41	0
021	Cobalt (ppm)	0553	0.2660	0.0020	0.3105	0.1267	0.0191	11	-0.35	021.53	0
021	Cobalt (ppm)	0939	0.2700	0.0200	0.3105	0.1267	0.0191	11	-0.32	021.53	0
021	Cobalt (ppm)	0407	0.2794	0.0042	0.3105	0.1267	0.0191	11	-0.25	021.53	0
021	Cobalt (ppm)	0510	0.2900	0.0400	0.3105	0.1267	0.0191	11	-0.16	021.43	0
021	Cobalt (ppm)	0407	0.2920	0.0075	0.3105	0.1267	0.0191	11	-0.15	021.43	0
021	Cobalt (ppm)	0407	0.4069	0.0193	0.3105	0.1267	0.0191	11	0.76	021.41	0
021	Cobalt (ppm)	0045	0.5155	0.0050	0.3105	0.1267	0.0191	11	1.62	021.42	0
021	Cobalt (ppm)	0169	0.7250	0.0300	0.3105	0.1267	0.0191	11	3.27	021.43	0
021	Cobalt (ppm)	0353	0.7100	0.9400	0.3105	0.1267	0.0191	11	3.15	021.43	1
021	Cobalt (ppm)	0186	< 0.5		0.3105	0.1267	0.0191	11		021.52	5
021	Cobalt (ppm)	0164	< 0.75		0.3105	0.1267	0.0191	11		021.31	5
021	Cobalt (ppm)	0682	< 1		0.3105	0.1267	0.0191	11		021.43	5
022	Copper (ppm)	0536	6.560	0.3400	14.25	1.689	0.6394	77	-4.56	022.31	0
022	Copper (ppm)	0598	11.27	0.0001	14.25	1.689	0.6394	77	-1.77	022.41	0
022	Copper (ppm)	0653	11.34	0.0500	14.25	1.689	0.6394	77	-1.73	022.41	0
022	Copper (ppm)	0123	11.40	0.4000	14.25	1.689	0.6394	77	-1.69	022.41	0
022	Copper (ppm)	0511	11.50	1.000	14.25	1.689	0.6394	77	-1.63	022.43	0
022	Copper (ppm)	0242	11.50	1.000	14.25	1.689	0.6394	77	-1.63	022.99	0
022	Copper (ppm)	0186	12.00	0.0000	14.25	1.689	0.6394	77	-1.34	022.42	0
022	Copper (ppm)	0520	12.17	0.2760	14.25	1.689	0.6394	77	-1.23	022.41	0
022	Copper (ppm)	0968	12.45	0.1050	14.25	1.689	0.6394	77	-1.07	022.43	0
022	Copper (ppm)	0848	12.50	0.1700	14.25	1.689	0.6394	77	-1.04	022.41	0
022	Copper (ppm)	0164	12.50	1.000	14.25	1.689	0.6394	77	-1.04	022.44	0
022	Copper (ppm)	0100	12.50	1.000	14.25	1.689	0.6394	77	-1.04	022.99	0
022	Copper (ppm)	0870	12.66	0.3500	14.25	1.689	0.6394	77	-0.95	022.42	0
022	Copper (ppm)	0407	12.73	0.0300	14.25	1.689	0.6394	77	-0.90	022.41	0
022	Copper (ppm)	0229	13.10	1.470	14.25	1.689	0.6394	77	-0.69	022.41	0
022	Copper (ppm)	2129	13.15	0.3000	14.25	1.689	0.6394	77	-0.65	022.42	0
022	Copper (ppm)	0263	13.37	0.0930	14.25	1.689	0.6394	77	-0.52	022.42	0
022	Copper (ppm)	0726	13.38	0.0800	14.25	1.689	0.6394	77	-0.52	022.42	0
022	Copper (ppm)	0345	13.45	0.5000	14.25	1.689	0.6394	77	-0.48	022.43	0
022	Copper (ppm)	0425	13.45	0.2600	14.25	1.689	0.6394	77	-0.48	022.43	0
022	Copper (ppm)	0910	13.50	1.000	14.25	1.689	0.6394	77	-0.45	022.41	0
022	Copper (ppm)	2319	13.50	0.2000	14.25	1.689	0.6394	77	-0.45	022.42	0
022	Copper (ppm)	0297	13.50	1.000	14.25	1.689	0.6394	77	-0.45	022.43	0
022	Copper (ppm)	0033	13.58	0.1600	14.25	1.689	0.6394	77	-0.40	022.43	0
022	Copper (ppm)	0037	13.60	0.4000	14.25	1.689	0.6394	77	-0.39	022.42	0
022	Copper (ppm)	0918	13.66	0.3500	14.25	1.689	0.6394	77	-0.36	022.43	0
022	Copper (ppm)	0407	13.66	0.0382	14.25	1.689	0.6394	77	-0.35	022.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
022	Copper (ppm)	0045	13.70	0.4000	14.25	1.689	0.6394	77	-0.33	022.42	0
022	Copper (ppm)	0009	13.85	0.3000	14.25	1.689	0.6394	77	-0.24	022.42	0
022	Copper (ppm)	0692	13.90	0.4000	14.25	1.689	0.6394	77	-0.21	022.42	0
022	Copper (ppm)	0265	13.90	0.2000	14.25	1.689	0.6394	77	-0.21	022.42	0
022	Copper (ppm)	0939	13.94	1.280	14.25	1.689	0.6394	77	-0.19	022.53	0
022	Copper (ppm)	0572	13.95	0.9000	14.25	1.689	0.6394	77	-0.18	022.53	0
022	Copper (ppm)	0098	13.96	0.0700	14.25	1.689	0.6394	77	-0.18	022.43	0
022	Copper (ppm)	0004	13.98	0.1500	14.25	1.689	0.6394	77	-0.17	022.41	0
022	Copper (ppm)	0035	14.00	0.0000	14.25	1.689	0.6394	77	-0.15	022.42	0
022	Copper (ppm)	0366	14.00	0.0000	14.25	1.689	0.6394	77	-0.15	022.42	0
022	Copper (ppm)	0035	14.00	0.0000	14.25	1.689	0.6394	77	-0.15	022.43	0
022	Copper (ppm)	0083	14.00	0.0000	14.25	1.689	0.6394	77	-0.15	022.43	0
022	Copper (ppm)	0208	14.01	0.1000	14.25	1.689	0.6394	77	-0.15	022.41	0
022	Copper (ppm)	0357	14.09	0.0460	14.25	1.689	0.6394	77	-0.10	022.42	0
022	Copper (ppm)	0553	14.10	0.4000	14.25	1.689	0.6394	77	-0.09	022.53	0
022	Copper (ppm)	0674	14.24	0.0000	14.25	1.689	0.6394	77	-0.01	022.41	0
022	Copper (ppm)	0278	14.25	0.5000	14.25	1.689	0.6394	77	0.00	022.42	0
022	Copper (ppm)	0019	14.26	0.3000	14.25	1.689	0.6394	77	0.00	022.41	0
022	Copper (ppm)	0010	14.30	0.8000	14.25	1.689	0.6394	77	0.03	022.33	0
022	Copper (ppm)	0027	14.41	0.1690	14.25	1.689	0.6394	77	0.09	022.43	0
022	Copper (ppm)	0693	14.48	0.7790	14.25	1.689	0.6394	77	0.14	022.42	0
022	Copper (ppm)	0202	14.50	0.2000	14.25	1.689	0.6394	77	0.15	022.43	0
022	Copper (ppm)	0186	14.50	1.000	14.25	1.689	0.6394	77	0.15	022.52	0
022	Copper (ppm)	2302	14.50	1.000	14.25	1.689	0.6394	77	0.15	022.99	0
022	Copper (ppm)	0066	14.57	0.6070	14.25	1.689	0.6394	77	0.19	022.44	0
022	Copper (ppm)	0098	14.60	0.4000	14.25	1.689	0.6394	77	0.20	022.41	0
022	Copper (ppm)	0682	14.80	0.0000	14.25	1.689	0.6394	77	0.32	022.43	0
022	Copper (ppm)	0049	14.89	0.7900	14.25	1.689	0.6394	77	0.37	022.41	0
022	Copper (ppm)	2326	14.90	0.6600	14.25	1.689	0.6394	77	0.38	022.43	0
022	Copper (ppm)	0407	14.96	1.645	14.25	1.689	0.6394	77	0.42	022.53	0
022	Copper (ppm)	0921	14.96	0.9600	14.25	1.689	0.6394	77	0.42	022.31	0
022	Copper (ppm)	0190	15.11	0.0800	14.25	1.689	0.6394	77	0.51	022.42	0
022	Copper (ppm)	0870	15.18	0.2700	14.25	1.689	0.6394	77	0.54	022.43	0
022	Copper (ppm)	0964	15.30	2.000	14.25	1.689	0.6394	77	0.62	022.43	0
022	Copper (ppm)	0169	15.35	0.1000	14.25	1.689	0.6394	77	0.65	022.43	0
022	Copper (ppm)	0510	15.50	1.000	14.25	1.689	0.6394	77	0.74	022.43	0
022	Copper (ppm)	0685	16.24	0.4000	14.25	1.689	0.6394	77	1.18	022.42	0
022	Copper (ppm)	0148	16.33	2.206	14.25	1.689	0.6394	77	1.23	022.41	0
022	Copper (ppm)	0529	16.75	1.500	14.25	1.689	0.6394	77	1.48	022.31	0
022	Copper (ppm)	0226	16.88	0.3300	14.25	1.689	0.6394	77	1.55	022.41	0
022	Copper (ppm)	0042	17.15	0.1000	14.25	1.689	0.6394	77	1.71	022.42	0
022	Copper (ppm)	0505	17.43	1.040	14.25	1.689	0.6394	77	1.88	022.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
022	Copper (ppm)	0171	17.66	0.9300	14.25	1.689	0.6394	77	2.01	022.41	0
022	Copper (ppm)	2188	17.70	0.3500	14.25	1.689	0.6394	77	2.04	022.31	0
022	Copper (ppm)	0353	18.00	2.000	14.25	1.689	0.6394	77	2.22	022.43	0
022	Copper (ppm)	2089	19.43	1.620	14.25	1.689	0.6394	77	3.06	022.43	0
022	Copper (ppm)	0047	20.93	2.720	14.25	1.689	0.6394	77	3.95	022.52	0
022	Copper (ppm)	0651	21.05	0.1302	14.25	1.689	0.6394	77	4.02	022.31	0
022	Copper (ppm)	0354	23.91	1.230	14.25	1.689	0.6394	77	5.71	022.41	0
022	Copper (ppm)	0011	60.00	5.600	14.25	1.689	0.6394	77	27.09	022.43	0
022	Copper (ppm)	2097	205.5	13.00	14.25	1.689	0.6394	77	113.24	022.31	2
025	Iron (ppm)	0037	45.80	10.80	91.43	7.794	2.322	70	-5.86	025.42	0
025	Iron (ppm)	0505	63.96	6.660	91.43	7.794	2.322	70	-3.53	025.43	0
025	Iron (ppm)	0651	69.70	2.980	91.43	7.794	2.322	70	-2.79	025.31	0
025	Iron (ppm)	0154	76.63	3.071	91.43	7.794	2.322	70	-1.90	025.52	0
025	Iron (ppm)	0511	77.50	1.000	91.43	7.794	2.322	70	-1.79	025.43	0
025	Iron (ppm)	0123	78.60	9.400	91.43	7.794	2.322	70	-1.65	025.41	0
025	Iron (ppm)	0297	79.00	0.0000	91.43	7.794	2.322	70	-1.60	025.43	0
025	Iron (ppm)	0598	80.26	1.065	91.43	7.794	2.322	70	-1.43	025.41	0
025	Iron (ppm)	0242	81.50	1.000	91.43	7.794	2.322	70	-1.27	025.99	0
025	Iron (ppm)	0682	82.81	0.0000	91.43	7.794	2.322	70	-1.11	025.43	0
025	Iron (ppm)	0265	83.65	0.9000	91.43	7.794	2.322	70	-1.00	025.42	0
025	Iron (ppm)	0910	84.00	4.000	91.43	7.794	2.322	70	-0.95	025.41	0
025	Iron (ppm)	0968	84.74	0.2930	91.43	7.794	2.322	70	-0.86	025.43	0
025	Iron (ppm)	0870	85.20	4.230	91.43	7.794	2.322	70	-0.80	025.42	0
025	Iron (ppm)	0066	85.58	0.9200	91.43	7.794	2.322	70	-0.75	025.41	0
025	Iron (ppm)	0653	86.32	0.0400	91.43	7.794	2.322	70	-0.66	025.41	0
025	Iron (ppm)	0876	86.80	2.600	91.43	7.794	2.322	70	-0.59	025.43	0
025	Iron (ppm)	0019	86.97	0.8300	91.43	7.794	2.322	70	-0.57	025.41	0
025	Iron (ppm)	0345	87.10	0.6000	91.43	7.794	2.322	70	-0.56	025.43	0
025	Iron (ppm)	0164	88.00	2.000	91.43	7.794	2.322	70	-0.44	025.41	0
025	Iron (ppm)	0918	88.59	0.5600	91.43	7.794	2.322	70	-0.36	025.43	0
025	Iron (ppm)	0921	88.65	5.070	91.43	7.794	2.322	70	-0.36	025.31	0
025	Iron (ppm)	2319	88.85	0.3000	91.43	7.794	2.322	70	-0.33	025.42	0
025	Iron (ppm)	0100	89.00	2.000	91.43	7.794	2.322	70	-0.31	025.99	0
025	Iron (ppm)	0425	89.32	0.7000	91.43	7.794	2.322	70	-0.27	025.43	0
025	Iron (ppm)	2188	89.46	0.2700	91.43	7.794	2.322	70	-0.25	025.31	0
025	Iron (ppm)	0263	89.57	0.0900	91.43	7.794	2.322	70	-0.24	025.42	0
025	Iron (ppm)	0563	89.66	1.974	91.43	7.794	2.322	70	-0.23	025.31	0
025	Iron (ppm)	0083	90.00	0.0000	91.43	7.794	2.322	70	-0.18	025.43	0
025	Iron (ppm)	0045	90.35	0.5000	91.43	7.794	2.322	70	-0.14	025.42	0
025	Iron (ppm)	0553	90.45	2.900	91.43	7.794	2.322	70	-0.13	025.53	0
025	Iron (ppm)	0278	90.50	3.000	91.43	7.794	2.322	70	-0.12	025.42	0
025	Iron (ppm)	0520	90.83	6.261	91.43	7.794	2.322	70	-0.08	025.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
025	Iron (ppm)	0208	90.83	1.240	91.43	7.794	2.322	70	-0.08	025.41	0
025	Iron (ppm)	0407	91.06	0.4770	91.43	7.794	2.322	70	-0.05	025.41	0
025	Iron (ppm)	0171	91.13	1.570	91.43	7.794	2.322	70	-0.04	025.41	0
025	Iron (ppm)	0870	91.28	1.720	91.43	7.794	2.322	70	-0.02	025.43	0
025	Iron (ppm)	0098	91.54	0.0800	91.43	7.794	2.322	70	0.01	025.41	0
025	Iron (ppm)	0033	92.03	0.5500	91.43	7.794	2.322	70	0.08	025.43	0
025	Iron (ppm)	0964	92.05	0.1000	91.43	7.794	2.322	70	0.08	025.43	0
025	Iron (ppm)	0226	92.51	2.080	91.43	7.794	2.322	70	0.14	025.41	0
025	Iron (ppm)	0693	92.54	1.142	91.43	7.794	2.322	70	0.14	025.42	0
025	Iron (ppm)	0004	93.00	1.430	91.43	7.794	2.322	70	0.20	025.41	0
025	Iron (ppm)	0027	93.02	0.4800	91.43	7.794	2.322	70	0.20	025.43	0
025	Iron (ppm)	0692	93.40	5.000	91.43	7.794	2.322	70	0.25	025.42	0
025	Iron (ppm)	0939	93.90	0.4300	91.43	7.794	2.322	70	0.32	025.53	0
025	Iron (ppm)	0169	94.20	1.800	91.43	7.794	2.322	70	0.35	025.43	0
025	Iron (ppm)	0098	94.21	0.1200	91.43	7.794	2.322	70	0.36	025.43	0
025	Iron (ppm)	0726	94.33	0.0500	91.43	7.794	2.322	70	0.37	025.42	0
025	Iron (ppm)	0035	95.00	10.00	91.43	7.794	2.322	70	0.46	025.42	0
025	Iron (ppm)	0148	95.26	3.055	91.43	7.794	2.322	70	0.49	025.41	0
025	Iron (ppm)	0013	95.70	1.400	91.43	7.794	2.322	70	0.55	025.43	0
025	Iron (ppm)	0529	96.70	0.6000	91.43	7.794	2.322	70	0.68	025.31	0
025	Iron (ppm)	0407	97.02	8.035	91.43	7.794	2.322	70	0.72	025.53	0
025	Iron (ppm)	0510	97.50	3.000	91.43	7.794	2.322	70	0.78	025.43	0
025	Iron (ppm)	0354	98.10	4.710	91.43	7.794	2.322	70	0.85	025.41	0
025	Iron (ppm)	0366	98.50	1.000	91.43	7.794	2.322	70	0.91	025.42	0
025	Iron (ppm)	0049	98.73	2.700	91.43	7.794	2.322	70	0.94	025.41	0
025	Iron (ppm)	0035	99.50	1.000	91.43	7.794	2.322	70	1.03	025.43	0
025	Iron (ppm)	0674	99.55	0.0000	91.43	7.794	2.322	70	1.04	025.41	0
025	Iron (ppm)	0407	100.2	1.838	91.43	7.794	2.322	70	1.13	025.43	0
025	Iron (ppm)	2302	100.5	1.000	91.43	7.794	2.322	70	1.16	025.99	0
025	Iron (ppm)	2326	101.0	2.000	91.43	7.794	2.322	70	1.23	025.43	0
025	Iron (ppm)	0190	101.1	1.400	91.43	7.794	2.322	70	1.24	025.42	0
025	Iron (ppm)	2192	103.0	2.000	91.43	7.794	2.322	70	1.48	025.43	0
025	Iron (ppm)	0042	111.0	6.000	91.43	7.794	2.322	70	2.51	025.42	0
025	Iron (ppm)	0536	111.9	2.230	91.43	7.794	2.322	70	2.63	025.31	0
025	Iron (ppm)	2089	135.8	1.470	91.43	7.794	2.322	70	5.70	025.43	0
025	Iron (ppm)	0047	136.7	5.800	91.43	7.794	2.322	70	5.81	025.52	0
025	Iron (ppm)	0353	159.5	9.000	91.43	7.794	2.322	70	8.73	025.43	0
025	Iron (ppm)	0229	97.47	14.47	91.43	7.794	2.322	70	0.77	025.41	1
025	Iron (ppm)	2097	152.5	19.00	91.43	7.794	2.322	70	7.84	025.31	1
025	Iron (ppm)	2129	517.4	9.300	91.43	7.794	2.322	70	54.65	025.42	2
027	Magnesium (%)	0529	0.2155	0.0070	0.2774	0.0139	0.0096	68	-4.46	027.31	0
027	Magnesium (%)	0407	0.2407	0.0052	0.2774	0.0139	0.0096	68	-2.64	027.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
027	Magnesium (%)	0407	0.2434	0.0025	0.2774	0.0139	0.0096	68	-2.45	027.43	0
027	Magnesium (%)	0425	0.2500	0.0200	0.2774	0.0139	0.0096	68	-1.97	027.43	0
027	Magnesium (%)	0968	0.2525	0.0030	0.2774	0.0139	0.0096	68	-1.79	027.43	0
027	Magnesium (%)	0910	0.2550	0.0100	0.2774	0.0139	0.0096	68	-1.61	027.41	0
027	Magnesium (%)	0278	0.2600	0.0000	0.2774	0.0139	0.0096	68	-1.25	027.42	0
027	Magnesium (%)	0083	0.2600	0.0000	0.2774	0.0139	0.0096	68	-1.25	027.43	0
027	Magnesium (%)	0511	0.2600	0.0200	0.2774	0.0139	0.0096	68	-1.25	027.43	0
027	Magnesium (%)	0242	0.2600	0.0000	0.2774	0.0139	0.0096	68	-1.25	027.99	0
027	Magnesium (%)	0263	0.2620	0.0080	0.2774	0.0139	0.0096	68	-1.11	027.42	0
027	Magnesium (%)	0520	0.2640	0.0300	0.2774	0.0139	0.0096	68	-0.96	027.41	0
027	Magnesium (%)	0265	0.2650	0.0100	0.2774	0.0139	0.0096	68	-0.89	027.42	0
027	Magnesium (%)	0870	0.2677	0.0041	0.2774	0.0139	0.0096	68	-0.70	027.43	0
027	Magnesium (%)	0870	0.2680	0.0044	0.2774	0.0139	0.0096	68	-0.68	027.42	0
027	Magnesium (%)	0045	0.2690	0.0120	0.2774	0.0139	0.0096	68	-0.60	027.42	0
027	Magnesium (%)	0229	0.2700	0.0400	0.2774	0.0139	0.0096	68	-0.53	027.41	0
027	Magnesium (%)	2319	0.2700	0.0000	0.2774	0.0139	0.0096	68	-0.53	027.42	0
027	Magnesium (%)	0164	0.2700	0.0080	0.2774	0.0139	0.0096	68	-0.53	027.44	0
027	Magnesium (%)	0300	0.2703	0.0372	0.2774	0.0139	0.0096	68	-0.51	027.43	0
027	Magnesium (%)	0186	0.2710	0.0100	0.2774	0.0139	0.0096	68	-0.46	027.52	0
027	Magnesium (%)	0142	0.2711	0.0052	0.2774	0.0139	0.0096	68	-0.45	027.31	0
027	Magnesium (%)	0010	0.2750	0.0100	0.2774	0.0139	0.0096	68	-0.17	027.33	0
027	Magnesium (%)	0049	0.2750	0.0100	0.2774	0.0139	0.0096	68	-0.17	027.41	0
027	Magnesium (%)	0123	0.2750	0.0100	0.2774	0.0139	0.0096	68	-0.17	027.41	0
027	Magnesium (%)	0685	0.2750	0.0100	0.2774	0.0139	0.0096	68	-0.17	027.42	0
027	Magnesium (%)	0202	0.2750	0.0100	0.2774	0.0139	0.0096	68	-0.17	027.43	0
027	Magnesium (%)	0297	0.2750	0.0300	0.2774	0.0139	0.0096	68	-0.17	027.43	0
027	Magnesium (%)	0353	0.2750	0.0100	0.2774	0.0139	0.0096	68	-0.17	027.43	0
027	Magnesium (%)	2192	0.2750	0.0100	0.2774	0.0139	0.0096	68	-0.17	027.43	0
027	Magnesium (%)	0939	0.2750	0.0300	0.2774	0.0139	0.0096	68	-0.17	027.53	0
027	Magnesium (%)	0598	0.2752	0.0006	0.2774	0.0139	0.0096	68	-0.16	027.41	0
027	Magnesium (%)	0345	0.2755	0.0050	0.2774	0.0139	0.0096	68	-0.14	027.43	0
027	Magnesium (%)	0553	0.2760	0.0200	0.2774	0.0139	0.0096	68	-0.10	027.53	0
027	Magnesium (%)	0186	0.2765	0.0130	0.2774	0.0139	0.0096	68	-0.06	027.42	0
027	Magnesium (%)	0033	0.2765	0.0010	0.2774	0.0139	0.0096	68	-0.06	027.43	0
027	Magnesium (%)	0037	0.2770	0.0060	0.2774	0.0139	0.0096	68	-0.03	027.42	0
027	Magnesium (%)	0148	0.2780	0.0100	0.2774	0.0139	0.0096	68	0.04	027.41	0
027	Magnesium (%)	0407	0.2797	0.0106	0.2774	0.0139	0.0096	68	0.17	027.53	0
027	Magnesium (%)	0004	0.2800	0.0200	0.2774	0.0139	0.0096	68	0.19	027.41	0
027	Magnesium (%)	0098	0.2800	0.0000	0.2774	0.0139	0.0096	68	0.19	027.41	0
027	Magnesium (%)	0692	0.2800	0.0200	0.2774	0.0139	0.0096	68	0.19	027.42	0
027	Magnesium (%)	0726	0.2800	0.0000	0.2774	0.0139	0.0096	68	0.19	027.42	0
027	Magnesium (%)	0098	0.2800	0.0000	0.2774	0.0139	0.0096	68	0.19	027.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
027	Magnesium (%)	2089	0.2800	0.0000	0.2774	0.0139	0.0096	68	0.19	027.43	0
027	Magnesium (%)	0357	0.2825	0.0001	0.2774	0.0139	0.0096	68	0.37	027.42	0
027	Magnesium (%)	0918	0.2845	0.0030	0.2774	0.0139	0.0096	68	0.51	027.43	0
027	Magnesium (%)	0226	0.2850	0.0100	0.2774	0.0139	0.0096	68	0.55	027.41	0
027	Magnesium (%)	0035	0.2850	0.0100	0.2774	0.0139	0.0096	68	0.55	027.42	0
027	Magnesium (%)	0366	0.2850	0.0100	0.2774	0.0139	0.0096	68	0.55	027.42	0
027	Magnesium (%)	0572	0.2850	0.0220	0.2774	0.0139	0.0096	68	0.55	027.53	0
027	Magnesium (%)	0653	0.2855	0.0010	0.2774	0.0139	0.0096	68	0.59	027.41	0
027	Magnesium (%)	0154	0.2870	0.0009	0.2774	0.0139	0.0096	68	0.69	027.52	0
027	Magnesium (%)	0563	0.2873	0.0036	0.2774	0.0139	0.0096	68	0.72	027.31	0
027	Magnesium (%)	0354	0.2900	0.0000	0.2774	0.0139	0.0096	68	0.91	027.41	0
027	Magnesium (%)	0035	0.2900	0.0000	0.2774	0.0139	0.0096	68	0.91	027.43	0
027	Magnesium (%)	0100	0.2900	0.0000	0.2774	0.0139	0.0096	68	0.91	027.99	0
027	Magnesium (%)	0693	0.2915	0.0230	0.2774	0.0139	0.0096	68	1.02	027.42	0
027	Magnesium (%)	0876	0.2925	0.0230	0.2774	0.0139	0.0096	68	1.09	027.43	0
027	Magnesium (%)	0536	0.2950	0.0100	0.2774	0.0139	0.0096	68	1.27	027.31	0
027	Magnesium (%)	0190	0.2950	0.0100	0.2774	0.0139	0.0096	68	1.27	027.42	0
027	Magnesium (%)	0510	0.2950	0.0100	0.2774	0.0139	0.0096	68	1.27	027.43	0
027	Magnesium (%)	0964	0.2998	0.0024	0.2774	0.0139	0.0096	68	1.62	027.43	0
027	Magnesium (%)	0682	0.3000	0.0000	0.2774	0.0139	0.0096	68	1.63	027.43	0
027	Magnesium (%)	0171	0.3050	0.0100	0.2774	0.0139	0.0096	68	1.99	027.41	0
027	Magnesium (%)	2326	0.3050	0.0100	0.2774	0.0139	0.0096	68	1.99	027.43	0
027	Magnesium (%)	0042	0.3095	0.0030	0.2774	0.0139	0.0096	68	2.32	027.42	0
027	Magnesium (%)	0169	0.3100	0.0200	0.2774	0.0139	0.0096	68	2.35	027.32	0
027	Magnesium (%)	2129	0.7541	0.0032	0.2774	0.0139	0.0096	68	34.37	027.42	2
028	Manganese (ppm)	0536	22.21	0.7100	41.56	3.824	1.252	78	-5.06	028.31	0
028	Manganese (ppm)	0186	32.00	2.000	41.56	3.824	1.252	78	-2.50	028.42	0
028	Manganese (ppm)	0190	35.25	0.0200	41.56	3.824	1.252	78	-1.65	028.42	0
028	Manganese (ppm)	0047	35.66	0.5600	41.56	3.824	1.252	78	-1.54	028.52	0
028	Manganese (ppm)	0066	35.79	0.6300	41.56	3.824	1.252	78	-1.51	028.44	0
028	Manganese (ppm)	0870	36.04	0.5200	41.56	3.824	1.252	78	-1.44	028.42	0
028	Manganese (ppm)	0123	36.25	3.900	41.56	3.824	1.252	78	-1.39	028.41	0
028	Manganese (ppm)	0870	36.34	2.290	41.56	3.824	1.252	78	-1.37	028.43	0
028	Manganese (ppm)	0229	36.37	4.740	41.56	3.824	1.252	78	-1.36	028.41	0
028	Manganese (ppm)	0511	36.50	1.000	41.56	3.824	1.252	78	-1.32	028.43	0
028	Manganese (ppm)	0598	37.56	0.2081	41.56	3.824	1.252	78	-1.04	028.41	0
028	Manganese (ppm)	0563	37.75	0.6700	41.56	3.824	1.252	78	-1.00	028.31	0
028	Manganese (ppm)	0968	37.78	0.4120	41.56	3.824	1.252	78	-0.99	028.43	0
028	Manganese (ppm)	0685	38.05	0.6000	41.56	3.824	1.252	78	-0.92	028.42	0
028	Manganese (ppm)	0171	38.26	0.0300	41.56	3.824	1.252	78	-0.86	028.41	0
028	Manganese (ppm)	0651	38.84	0.7397	41.56	3.824	1.252	78	-0.71	028.31	0
028	Manganese (ppm)	0529	39.00	1.800	41.56	3.824	1.252	78	-0.67	028.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
028	Manganese (ppm)	0910	39.00	2.000	41.56	3.824	1.252	78	-0.67	028.41	0
028	Manganese (ppm)	0009	39.00	0.0000	41.56	3.824	1.252	78	-0.67	028.42	0
028	Manganese (ppm)	0278	39.00	2.000	41.56	3.824	1.252	78	-0.67	028.42	0
028	Manganese (ppm)	0208	39.10	3.000	41.56	3.824	1.252	78	-0.64	028.31	0
028	Manganese (ppm)	0921	39.13	2.240	41.56	3.824	1.252	78	-0.64	028.31	0
028	Manganese (ppm)	0263	39.29	0.4870	41.56	3.824	1.252	78	-0.59	028.42	0
028	Manganese (ppm)	0242	39.50	1.000	41.56	3.824	1.252	78	-0.54	028.99	0
028	Manganese (ppm)	0653	39.78	0.1600	41.56	3.824	1.252	78	-0.47	028.41	0
028	Manganese (ppm)	0553	39.80	2.000	41.56	3.824	1.252	78	-0.46	028.53	0
028	Manganese (ppm)	0520	39.95	1.477	41.56	3.824	1.252	78	-0.42	028.41	0
028	Manganese (ppm)	0692	39.95	4.700	41.56	3.824	1.252	78	-0.42	028.42	0
028	Manganese (ppm)	0297	40.00	4.000	41.56	3.824	1.252	78	-0.41	028.43	0
028	Manganese (ppm)	0164	40.00	0.0000	41.56	3.824	1.252	78	-0.41	028.44	0
028	Manganese (ppm)	0033	40.36	0.3200	41.56	3.824	1.252	78	-0.31	028.43	0
028	Manganese (ppm)	0208	40.65	0.4800	41.56	3.824	1.252	78	-0.24	028.41	0
028	Manganese (ppm)	2319	40.75	0.7000	41.56	3.824	1.252	78	-0.21	028.42	0
028	Manganese (ppm)	0354	41.05	0.0000	41.56	3.824	1.252	78	-0.13	028.41	0
028	Manganese (ppm)	0098	41.32	0.3300	41.56	3.824	1.252	78	-0.06	028.41	0
028	Manganese (ppm)	0049	41.51	1.390	41.56	3.824	1.252	78	-0.01	028.41	0
028	Manganese (ppm)	0407	41.53	1.030	41.56	3.824	1.252	78	-0.01	028.41	0
028	Manganese (ppm)	0572	41.60	0.2000	41.56	3.824	1.252	78	0.01	028.53	0
028	Manganese (ppm)	0918	41.69	1.230	41.56	3.824	1.252	78	0.03	028.43	0
028	Manganese (ppm)	0345	41.75	0.5000	41.56	3.824	1.252	78	0.05	028.43	0
028	Manganese (ppm)	0083	42.00	0.0000	41.56	3.824	1.252	78	0.12	028.43	0
028	Manganese (ppm)	0510	42.00	0.0000	41.56	3.824	1.252	78	0.12	028.43	0
028	Manganese (ppm)	0045	42.05	1.700	41.56	3.824	1.252	78	0.13	028.42	0
028	Manganese (ppm)	0019	42.17	0.2400	41.56	3.824	1.252	78	0.16	028.41	0
028	Manganese (ppm)	0169	42.25	1.300	41.56	3.824	1.252	78	0.18	028.43	0
028	Manganese (ppm)	0098	42.37	0.5400	41.56	3.824	1.252	78	0.21	028.43	0
028	Manganese (ppm)	0100	42.50	1.000	41.56	3.824	1.252	78	0.25	028.99	0
028	Manganese (ppm)	0004	42.52	1.270	41.56	3.824	1.252	78	0.25	028.41	0
028	Manganese (ppm)	0693	42.62	1.500	41.56	3.824	1.252	78	0.28	028.42	0
028	Manganese (ppm)	2188	42.66	0.3900	41.56	3.824	1.252	78	0.29	028.31	0
028	Manganese (ppm)	0186	43.00	4.000	41.56	3.824	1.252	78	0.38	028.52	0
028	Manganese (ppm)	0425	43.14	0.1100	41.56	3.824	1.252	78	0.41	028.43	0
028	Manganese (ppm)	0013	43.20	4.400	41.56	3.824	1.252	78	0.43	028.43	0
028	Manganese (ppm)	0674	43.28	0.0000	41.56	3.824	1.252	78	0.45	028.41	0
028	Manganese (ppm)	0148	43.33	1.260	41.56	3.824	1.252	78	0.46	028.41	0
028	Manganese (ppm)	0027	43.69	0.2310	41.56	3.824	1.252	78	0.56	028.43	0
028	Manganese (ppm)	0226	44.01	1.450	41.56	3.824	1.252	78	0.64	028.41	0
028	Manganese (ppm)	0407	44.05	0.7991	41.56	3.824	1.252	78	0.65	028.43	0
028	Manganese (ppm)	2192	44.20	2.200	41.56	3.824	1.252	78	0.69	028.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
028	Manganese (ppm)	0357	44.37	0.7360	41.56	3.824	1.252	78	0.74	028.42	0
028	Manganese (ppm)	0939	44.37	2.660	41.56	3.824	1.252	78	0.74	028.53	0
028	Manganese (ppm)	0035	44.50	1.000	41.56	3.824	1.252	78	0.77	028.43	0
028	Manganese (ppm)	0011	44.95	0.3000	41.56	3.824	1.252	78	0.89	028.43	0
028	Manganese (ppm)	0035	45.00	0.0000	41.56	3.824	1.252	78	0.90	028.42	0
028	Manganese (ppm)	0964	45.40	0.4000	41.56	3.824	1.252	78	1.00	028.43	0
028	Manganese (ppm)	2089	45.46	2.050	41.56	3.824	1.252	78	1.02	028.43	0
028	Manganese (ppm)	2326	45.50	3.000	41.56	3.824	1.252	78	1.03	028.43	0
028	Manganese (ppm)	2302	45.50	1.000	41.56	3.824	1.252	78	1.03	028.99	0
028	Manganese (ppm)	0037	45.55	1.700	41.56	3.824	1.252	78	1.04	028.42	0
028	Manganese (ppm)	0265	46.05	1.500	41.56	3.824	1.252	78	1.17	028.42	0
028	Manganese (ppm)	0407	46.41	0.6036	41.56	3.824	1.252	78	1.27	028.53	0
028	Manganese (ppm)	0202	46.55	1.500	41.56	3.824	1.252	78	1.31	028.43	0
028	Manganese (ppm)	0726	46.96	0.8700	41.56	3.824	1.252	78	1.41	028.42	0
028	Manganese (ppm)	0366	47.50	1.000	41.56	3.824	1.252	78	1.55	028.42	0
028	Manganese (ppm)	0505	47.75	0.5000	41.56	3.824	1.252	78	1.62	028.43	0
028	Manganese (ppm)	0042	48.00	1.400	41.56	3.824	1.252	78	1.68	028.42	0
028	Manganese (ppm)	0353	50.50	5.000	41.56	3.824	1.252	78	2.34	028.43	0
028	Manganese (ppm)	0682	53.65	0.0000	41.56	3.824	1.252	78	3.16	028.43	0
028	Manganese (ppm)	2097	127.5	43.00	41.56	3.824	1.252	78	22.47	028.31	1
028	Manganese (ppm)	2129	425.0	4.000	41.56	3.824	1.252	78	100.26	028.42	2
031	Phosphorus (%)	2192	0.2900	0.0000	0.6647	0.0287	0.0183	89	-13.07	031.01	0
031	Phosphorus (%)	2022	0.5850	0.0100	0.6647	0.0287	0.0183	89	-2.78	031.01	0
031	Phosphorus (%)	0229	0.5950	0.0700	0.6647	0.0287	0.0183	89	-2.43	031.41	0
031	Phosphorus (%)	0186	0.6034	0.0263	0.6647	0.0287	0.0183	89	-2.14	031.42	0
031	Phosphorus (%)	0511	0.6050	0.0300	0.6647	0.0287	0.0183	89	-2.08	031.43	0
031	Phosphorus (%)	2006	0.6055	0.0290	0.6647	0.0287	0.0183	89	-2.07	031.01	0
031	Phosphorus (%)	0968	0.6075	0.0110	0.6647	0.0287	0.0183	89	-2.00	031.43	0
031	Phosphorus (%)	0242	0.6100	0.0200	0.6647	0.0287	0.0183	89	-1.91	031.99	0
031	Phosphorus (%)	0626	0.6125	0.0110	0.6647	0.0287	0.0183	89	-1.82	031.01	0
031	Phosphorus (%)	0687	0.6250	0.0500	0.6647	0.0287	0.0183	89	-1.39	031.01	0
031	Phosphorus (%)	0676	0.6250	0.0100	0.6647	0.0287	0.0183	89	-1.39	031.99	0
031	Phosphorus (%)	0263	0.6290	0.0120	0.6647	0.0287	0.0183	89	-1.25	031.42	0
031	Phosphorus (%)	2193	0.6300	0.0200	0.6647	0.0287	0.0183	89	-1.21	031.01	0
031	Phosphorus (%)	0100	0.6300	0.0000	0.6647	0.0287	0.0183	89	-1.21	031.99	0
031	Phosphorus (%)	0164	0.6380	0.0140	0.6647	0.0287	0.0183	89	-0.93	031.44	0
031	Phosphorus (%)	0425	0.6400	0.0000	0.6647	0.0287	0.0183	89	-0.86	031.43	0
031	Phosphorus (%)	0682	0.6400	0.0000	0.6647	0.0287	0.0183	89	-0.86	031.43	0
031	Phosphorus (%)	2290	0.6400	0.0000	0.6647	0.0287	0.0183	89	-0.86	031.99	0
031	Phosphorus (%)	0142	0.6429	0.0122	0.6647	0.0287	0.0183	89	-0.76	031.01	0
031	Phosphorus (%)	0563	0.6436	0.0220	0.6647	0.0287	0.0183	89	-0.74	031.01	0
031	Phosphorus (%)	0536	0.6450	0.0100	0.6647	0.0287	0.0183	89	-0.69	031.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	0910	0.6450	0.0300	0.6647	0.0287	0.0183	89	-0.69	031.41	0
031	Phosphorus (%)	2188	0.6500	0.0000	0.6647	0.0287	0.0183	89	-0.51	031.01	0
031	Phosphorus (%)	0876	0.6510	0.0340	0.6647	0.0287	0.0183	89	-0.48	031.43	0
031	Phosphorus (%)	0553	0.6535	0.0990	0.6647	0.0287	0.0183	89	-0.39	031.53	0
031	Phosphorus (%)	0407	0.6547	0.0110	0.6647	0.0287	0.0183	89	-0.35	031.41	0
031	Phosphorus (%)	0004	0.6550	0.0100	0.6647	0.0287	0.0183	89	-0.34	031.41	0
031	Phosphorus (%)	0354	0.6550	0.0100	0.6647	0.0287	0.0183	89	-0.34	031.41	0
031	Phosphorus (%)	0685	0.6550	0.0100	0.6647	0.0287	0.0183	89	-0.34	031.41	0
031	Phosphorus (%)	2322	0.6550	0.0100	0.6647	0.0287	0.0183	89	-0.34	031.43	0
031	Phosphorus (%)	0693	0.6565	0.0050	0.6647	0.0287	0.0183	89	-0.29	031.42	0
031	Phosphorus (%)	0300	0.6567	0.0127	0.6647	0.0287	0.0183	89	-0.28	031.43	0
031	Phosphorus (%)	0186	0.6580	0.0245	0.6647	0.0287	0.0183	89	-0.24	031.52	0
031	Phosphorus (%)	0345	0.6599	0.0423	0.6647	0.0287	0.0183	89	-0.17	031.43	0
031	Phosphorus (%)	0366	0.6600	0.0600	0.6647	0.0287	0.0183	89	-0.17	031.42	0
031	Phosphorus (%)	0884	0.6600	0.0000	0.6647	0.0287	0.0183	89	-0.17	031.01	0
031	Phosphorus (%)	0171	0.6600	0.0000	0.6647	0.0287	0.0183	89	-0.17	031.41	0
031	Phosphorus (%)	0010	0.6600	0.0200	0.6647	0.0287	0.0183	89	-0.17	031.43	0
031	Phosphorus (%)	0033	0.6600	0.0060	0.6647	0.0287	0.0183	89	-0.17	031.43	0
031	Phosphorus (%)	0297	0.6600	0.0400	0.6647	0.0287	0.0183	89	-0.17	031.43	0
031	Phosphorus (%)	0653	0.6605	0.0030	0.6647	0.0287	0.0183	89	-0.15	031.41	0
031	Phosphorus (%)	0651	0.6637	0.0040	0.6647	0.0287	0.0183	89	-0.04	031.01	0
031	Phosphorus (%)	2259	0.6645	0.0210	0.6647	0.0287	0.0183	89	-0.01	031.41	0
031	Phosphorus (%)	2058	0.6650	0.0300	0.6647	0.0287	0.0183	89	0.01	031.01	0
031	Phosphorus (%)	0169	0.6650	0.0100	0.6647	0.0287	0.0183	89	0.01	031.03	0
031	Phosphorus (%)	0035	0.6650	0.0100	0.6647	0.0287	0.0183	89	0.01	031.42	0
031	Phosphorus (%)	0265	0.6650	0.0100	0.6647	0.0287	0.0183	89	0.01	031.42	0
031	Phosphorus (%)	0278	0.6650	0.0500	0.6647	0.0287	0.0183	89	0.01	031.42	0
031	Phosphorus (%)	0520	0.6695	0.0110	0.6647	0.0287	0.0183	89	0.17	031.41	0
031	Phosphorus (%)	0407	0.6695	0.0054	0.6647	0.0287	0.0183	89	0.17	031.43	0
031	Phosphorus (%)	2181	0.6700	0.0400	0.6647	0.0287	0.0183	89	0.18	031.01	0
031	Phosphorus (%)	0529	0.6700	0.0000	0.6647	0.0287	0.0183	89	0.18	031.01	0
031	Phosphorus (%)	0202	0.6700	0.0000	0.6647	0.0287	0.0183	89	0.18	031.43	0
031	Phosphorus (%)	0208	0.6710	0.0140	0.6647	0.0287	0.0183	89	0.22	031.41	0
031	Phosphorus (%)	2190	0.6722	0.0055	0.6647	0.0287	0.0183	89	0.26	031.01	0
031	Phosphorus (%)	0598	0.6745	0.0058	0.6647	0.0287	0.0183	89	0.34	031.41	0
031	Phosphorus (%)	0948	0.6750	0.0100	0.6647	0.0287	0.0183	89	0.36	031.01	0
031	Phosphorus (%)	0049	0.6750	0.0300	0.6647	0.0287	0.0183	89	0.36	031.41	0
031	Phosphorus (%)	0035	0.6750	0.0100	0.6647	0.0287	0.0183	89	0.36	031.43	0
031	Phosphorus (%)	0037	0.6780	0.0200	0.6647	0.0287	0.0183	89	0.46	031.42	0
031	Phosphorus (%)	2097	0.6781	0.0020	0.6647	0.0287	0.0183	89	0.47	031.01	0
031	Phosphorus (%)	0939	0.6800	0.0400	0.6647	0.0287	0.0183	89	0.53	031.01	0
031	Phosphorus (%)	0226	0.6800	0.0200	0.6647	0.0287	0.0183	89	0.53	031.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	0083	0.6800	0.0200	0.6647	0.0287	0.0183	89	0.53	031.43	0
031	Phosphorus (%)	0939	0.6800	0.0200	0.6647	0.0287	0.0183	89	0.53	031.53	0
031	Phosphorus (%)	0870	0.6808	0.0085	0.6647	0.0287	0.0183	89	0.56	031.42	0
031	Phosphorus (%)	0964	0.6816	0.0082	0.6647	0.0287	0.0183	89	0.59	031.43	0
031	Phosphorus (%)	2319	0.6820	0.0200	0.6647	0.0287	0.0183	89	0.60	031.42	0
031	Phosphorus (%)	0870	0.6830	0.0117	0.6647	0.0287	0.0183	89	0.64	031.43	0
031	Phosphorus (%)	2302	0.6850	0.0300	0.6647	0.0287	0.0183	89	0.71	031.01	0
031	Phosphorus (%)	0144	0.6850	0.0300	0.6647	0.0287	0.0183	89	0.71	031.41	0
031	Phosphorus (%)	0098	0.6854	0.0043	0.6647	0.0287	0.0183	89	0.72	031.41	0
031	Phosphorus (%)	0154	0.6856	0.0032	0.6647	0.0287	0.0183	89	0.73	031.52	0
031	Phosphorus (%)	0918	0.6860	0.0120	0.6647	0.0287	0.0183	89	0.74	031.43	0
031	Phosphorus (%)	0045	0.6865	0.0070	0.6647	0.0287	0.0183	89	0.76	031.42	0
031	Phosphorus (%)	0407	0.6907	0.0918	0.6647	0.0287	0.0183	89	0.91	031.53	0
031	Phosphorus (%)	0353	0.6950	0.0100	0.6647	0.0287	0.0183	89	1.06	031.43	0
031	Phosphorus (%)	2089	0.6950	0.0100	0.6647	0.0287	0.0183	89	1.06	031.43	0
031	Phosphorus (%)	0098	0.6996	0.0333	0.6647	0.0287	0.0183	89	1.21	031.43	0
031	Phosphorus (%)	2326	0.7000	0.0000	0.6647	0.0287	0.0183	89	1.23	031.43	0
031	Phosphorus (%)	0726	0.7100	0.0000	0.6647	0.0287	0.0183	89	1.58	031.42	0
031	Phosphorus (%)	0148	0.7160	0.0500	0.6647	0.0287	0.0183	89	1.79	031.41	0
031	Phosphorus (%)	0123	0.7250	0.0100	0.6647	0.0287	0.0183	89	2.10	031.41	0
031	Phosphorus (%)	0692	0.7250	0.0500	0.6647	0.0287	0.0183	89	2.10	031.42	0
031	Phosphorus (%)	0190	0.7300	0.0200	0.6647	0.0287	0.0183	89	2.28	031.42	0
031	Phosphorus (%)	0357	0.7314	0.0030	0.6647	0.0287	0.0183	89	2.33	031.42	0
031	Phosphorus (%)	0572	0.7345	0.0390	0.6647	0.0287	0.0183	89	2.43	031.53	0
031	Phosphorus (%)	0042	0.7485	0.0030	0.6647	0.0287	0.0183	89	2.92	031.42	0
031	Phosphorus (%)	0510	0.7500	0.0000	0.6647	0.0287	0.0183	89	2.97	031.43	0
031	Phosphorus (%)	2129	1.800	0.2000	0.6647	0.0287	0.0183	89	39.59	031.42	2
032	Potassium (%)	2129	1.225	0.0300	2.207	0.1300	0.0512	79	-7.56	032.42	0
032	Potassium (%)	0968	1.625	0.0070	2.207	0.1300	0.0512	79	-4.48	032.43	0
032	Potassium (%)	0169	1.875	0.0100	2.207	0.1300	0.0512	79	-2.56	032.32	0
032	Potassium (%)	0529	1.940	0.0800	2.207	0.1300	0.0512	79	-2.06	032.31	0
032	Potassium (%)	0366	1.945	0.0900	2.207	0.1300	0.0512	79	-2.02	032.42	0
032	Potassium (%)	0278	1.980	0.0000	2.207	0.1300	0.0512	79	-1.75	032.42	0
032	Potassium (%)	0563	1.982	0.0251	2.207	0.1300	0.0512	79	-1.73	032.31	0
032	Potassium (%)	0019	2.020	0.0000	2.207	0.1300	0.0512	79	-1.44	032.41	0
032	Potassium (%)	0297	2.035	0.0500	2.207	0.1300	0.0512	79	-1.32	032.43	0
032	Potassium (%)	0910	2.040	0.0600	2.207	0.1300	0.0512	79	-1.29	032.41	0
032	Potassium (%)	0425	2.040	0.1200	2.207	0.1300	0.0512	79	-1.29	032.43	0
032	Potassium (%)	0066	2.055	0.1100	2.207	0.1300	0.0512	79	-1.17	032.41	0
032	Potassium (%)	0354	2.090	0.0200	2.207	0.1300	0.0512	79	-0.90	032.41	0
032	Potassium (%)	0027	2.097	0.0300	2.207	0.1300	0.0512	79	-0.85	032.43	0
032	Potassium (%)	0511	2.100	0.1800	2.207	0.1300	0.0512	79	-0.82	032.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
032	Potassium (%)	0013	2.105	0.0500	2.207	0.1300	0.0512	79	-0.79	032.43	0
032	Potassium (%)	0918	2.110	0.0130	2.207	0.1300	0.0512	79	-0.75	032.43	0
032	Potassium (%)	0123	2.110	0.2200	2.207	0.1300	0.0512	79	-0.75	032.41	0
032	Potassium (%)	0083	2.115	0.0100	2.207	0.1300	0.0512	79	-0.71	032.43	0
032	Potassium (%)	0590	2.120	0.0200	2.207	0.1300	0.0512	79	-0.67	032.99	0
032	Potassium (%)	0100	2.125	0.0300	2.207	0.1300	0.0512	79	-0.63	032.99	0
032	Potassium (%)	0242	2.125	0.0100	2.207	0.1300	0.0512	79	-0.63	032.99	0
032	Potassium (%)	0598	2.138	0.0051	2.207	0.1300	0.0512	79	-0.54	032.41	0
032	Potassium (%)	0049	2.155	0.0100	2.207	0.1300	0.0512	79	-0.40	032.41	0
032	Potassium (%)	0505	2.160	0.0400	2.207	0.1300	0.0512	79	-0.36	032.43	0
032	Potassium (%)	0682	2.160	0.0000	2.207	0.1300	0.0512	79	-0.36	032.43	0
032	Potassium (%)	0171	2.165	0.0300	2.207	0.1300	0.0512	79	-0.32	032.41	0
032	Potassium (%)	0164	2.165	0.0700	2.207	0.1300	0.0512	79	-0.32	032.44	0
032	Potassium (%)	0004	2.170	0.0800	2.207	0.1300	0.0512	79	-0.29	032.41	0
032	Potassium (%)	0265	2.185	0.0700	2.207	0.1300	0.0512	79	-0.17	032.42	0
032	Potassium (%)	0353	2.195	0.0300	2.207	0.1300	0.0512	79	-0.09	032.43	0
032	Potassium (%)	2319	2.195	0.0300	2.207	0.1300	0.0512	79	-0.09	032.52	0
032	Potassium (%)	0693	2.196	0.0310	2.207	0.1300	0.0512	79	-0.09	032.42	0
032	Potassium (%)	0263	2.199	0.0180	2.207	0.1300	0.0512	79	-0.06	032.42	0
032	Potassium (%)	2188	2.200	0.0000	2.207	0.1300	0.0512	79	-0.05	032.08	0
032	Potassium (%)	0045	2.200	0.0200	2.207	0.1300	0.0512	79	-0.05	032.42	0
032	Potassium (%)	0047	2.200	0.0000	2.207	0.1300	0.0512	79	-0.05	032.42	0
032	Potassium (%)	0553	2.200	0.1400	2.207	0.1300	0.0512	79	-0.05	032.53	0
032	Potassium (%)	0154	2.201	0.0174	2.207	0.1300	0.0512	79	-0.04	032.52	0
032	Potassium (%)	0186	2.204	0.1190	2.207	0.1300	0.0512	79	-0.03	032.52	0
032	Potassium (%)	0186	2.206	0.1210	2.207	0.1300	0.0512	79	-0.01	032.42	0
032	Potassium (%)	2089	2.210	0.0200	2.207	0.1300	0.0512	79	0.02	032.43	0
032	Potassium (%)	0037	2.214	0.0440	2.207	0.1300	0.0512	79	0.05	032.42	0
032	Potassium (%)	0345	2.215	0.0100	2.207	0.1300	0.0512	79	0.06	032.43	0
032	Potassium (%)	0148	2.220	0.0200	2.207	0.1300	0.0512	79	0.10	032.41	0
032	Potassium (%)	0144	2.225	0.0700	2.207	0.1300	0.0512	79	0.14	032.41	0
032	Potassium (%)	0033	2.230	0.0400	2.207	0.1300	0.0512	79	0.18	032.43	0
032	Potassium (%)	0098	2.230	0.1000	2.207	0.1300	0.0512	79	0.18	032.43	0
032	Potassium (%)	0357	2.234	0.0426	2.207	0.1300	0.0512	79	0.20	032.42	0
032	Potassium (%)	0036	2.239	0.0555	2.207	0.1300	0.0512	79	0.25	032.43	0
032	Potassium (%)	0407	2.251	0.0697	2.207	0.1300	0.0512	79	0.34	032.41	0
032	Potassium (%)	0035	2.255	0.0100	2.207	0.1300	0.0512	79	0.37	032.42	0
032	Potassium (%)	0407	2.262	0.0140	2.207	0.1300	0.0512	79	0.42	032.43	0
032	Potassium (%)	0142	2.263	0.0566	2.207	0.1300	0.0512	79	0.43	032.31	0
032	Potassium (%)	0202	2.265	0.0700	2.207	0.1300	0.0512	79	0.45	032.43	0
032	Potassium (%)	0692	2.265	0.0500	2.207	0.1300	0.0512	79	0.45	032.42	0
032	Potassium (%)	0407	2.269	0.0722	2.207	0.1300	0.0512	79	0.47	032.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
032	Potassium (%)	0572	2.270	0.2000	2.207	0.1300	0.0512	79	0.48	032.53	0
032	Potassium (%)	0510	2.290	0.0200	2.207	0.1300	0.0512	79	0.64	032.43	0
032	Potassium (%)	0208	2.295	0.0300	2.207	0.1300	0.0512	79	0.68	032.31	0
032	Potassium (%)	0098	2.295	0.0100	2.207	0.1300	0.0512	79	0.68	032.41	0
032	Potassium (%)	0035	2.320	0.0000	2.207	0.1300	0.0512	79	0.87	032.43	0
032	Potassium (%)	0939	2.320	0.0800	2.207	0.1300	0.0512	79	0.87	032.53	0
032	Potassium (%)	0520	2.322	0.0560	2.207	0.1300	0.0512	79	0.88	032.41	0
032	Potassium (%)	0536	2.325	0.0700	2.207	0.1300	0.0512	79	0.91	032.31	0
032	Potassium (%)	0009	2.325	0.0900	2.207	0.1300	0.0512	79	0.91	032.42	0
032	Potassium (%)	2302	2.330	0.0000	2.207	0.1300	0.0512	79	0.95	032.99	0
032	Potassium (%)	0876	2.335	0.0500	2.207	0.1300	0.0512	79	0.98	032.43	0
032	Potassium (%)	0226	2.340	0.0200	2.207	0.1300	0.0512	79	1.02	032.41	0
032	Potassium (%)	0010	2.340	0.0800	2.207	0.1300	0.0512	79	1.02	032.43	0
032	Potassium (%)	2326	2.350	0.1000	2.207	0.1300	0.0512	79	1.10	032.43	0
032	Potassium (%)	0870	2.382	0.1240	2.207	0.1300	0.0512	79	1.35	032.42	0
032	Potassium (%)	0964	2.433	0.0251	2.207	0.1300	0.0512	79	1.74	032.43	0
032	Potassium (%)	0870	2.434	0.1190	2.207	0.1300	0.0512	79	1.74	032.43	0
032	Potassium (%)	0011	2.450	0.1000	2.207	0.1300	0.0512	79	1.87	032.43	0
032	Potassium (%)	0190	2.475	0.0300	2.207	0.1300	0.0512	79	2.06	032.42	0
032	Potassium (%)	0726	2.500	0.0200	2.207	0.1300	0.0512	79	2.25	032.42	0
032	Potassium (%)	0042	2.530	0.0200	2.207	0.1300	0.0512	79	2.48	032.42	0
032	Potassium (%)	0653	2.558	0.0600	2.207	0.1300	0.0512	79	2.70	032.41	0
032	Potassium (%)	0229	2.040	0.3200	2.207	0.1300	0.0512	79	-1.29	032.41	1
033	Salt as chloride (%)	0100	0.0100	0.0000	0.0653	0.0575	0.0140	27	-0.96	033.01	0
033	Salt as chloride (%)	0910	0.0100	0.0000	0.0653	0.0575	0.0140	27	-0.96	033.99	0
033	Salt as chloride (%)	0171	0.0150	0.0100	0.0653	0.0575	0.0140	27	-0.87	033.05	0
033	Salt as chloride (%)	0693	0.0160	0.0000	0.0653	0.0575	0.0140	27	-0.86	033.00	0
033	Salt as chloride (%)	0650	0.0200	0.0000	0.0653	0.0575	0.0140	27	-0.79	033.99	0
033	Salt as chloride (%)	0354	0.0225	0.0050	0.0653	0.0575	0.0140	27	-0.74	033.01	0
033	Salt as chloride (%)	0407	0.0245	0.0010	0.0653	0.0575	0.0140	27	-0.71	033.01	0
033	Salt as chloride (%)	0148	0.0289	0.0017	0.0653	0.0575	0.0140	27	-0.63	033.01	0
033	Salt as chloride (%)	0098	0.0300	0.0000	0.0653	0.0575	0.0140	27	-0.61	033.01	0
033	Salt as chloride (%)	0242	0.0300	0.0000	0.0653	0.0575	0.0140	27	-0.61	033.01	0
033	Salt as chloride (%)	0229	0.0350	0.0300	0.0653	0.0575	0.0140	27	-0.53	033.01	0
033	Salt as chloride (%)	0948	0.0350	0.0100	0.0653	0.0575	0.0140	27	-0.53	033.01	0
033	Salt as chloride (%)	0226	0.0400	0.0000	0.0653	0.0575	0.0140	27	-0.44	033.01	0
033	Salt as chloride (%)	2076	0.0475	0.0090	0.0653	0.0575	0.0140	27	-0.31	033.00	0
033	Salt as chloride (%)	0539	0.0500	0.0200	0.0653	0.0575	0.0140	27	-0.27	033.00	0
033	Salt as chloride (%)	2188	0.0500	0.0000	0.0653	0.0575	0.0140	27	-0.27	033.05	0
033	Salt as chloride (%)	0425	0.0600	0.0200	0.0653	0.0575	0.0140	27	-0.09	033.01	0
033	Salt as chloride (%)	0353	0.0650	0.0100	0.0653	0.0575	0.0140	27	-0.01	033.00	0
033	Salt as chloride (%)	0682	0.0900	0.0000	0.0653	0.0575	0.0140	27	0.43	033.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
033	Salt as chloride (%)	0144	0.1000	0.0200	0.0653	0.0575	0.0140	27	0.60	033.03	0
033	Salt as chloride (%)	2190	0.1051	0.0123	0.0653	0.0575	0.0140	27	0.69	033.00	0
033	Salt as chloride (%)	0510	0.1320	0.0000	0.0653	0.0575	0.0140	27	1.16	033.01	0
033	Salt as chloride (%)	0726	0.1400	0.0000	0.0653	0.0575	0.0140	27	1.30	033.03	0
033	Salt as chloride (%)	0366	0.1600	0.0000	0.0653	0.0575	0.0140	27	1.65	033.00	0
033	Salt as chloride (%)	2268	0.2450	0.0900	0.0653	0.0575	0.0140	27	3.12	033.01	0
033	Salt as chloride (%)	0265	0.2750	0.0100	0.0653	0.0575	0.0140	27	3.64	033.03	0
033	Salt as chloride (%)	0278	2.315	0.1300	0.0653	0.0575	0.0140	27	39.09	033.01	0
033	Salt as chloride (%)	2181	5.585	0.2300	0.0653	0.0575	0.0140	27	95.92	033.99	2
033	Salt as chloride (%)	0083	0.0000	0.0000	0.0653	0.0575	0.0140	27		033.01	4
033	Salt as chloride (%)	2043	0.0000	0.0000	0.0653	0.0575	0.0140	27		033.05	4
033	Salt as chloride (%)	2073	< 0.05		0.0653	0.0575	0.0140	27		033.99	5
033	Salt as chloride (%)	0309	< 0.1		0.0653	0.0575	0.0140	27		033.00	5
033	Salt as chloride (%)	0164	< 0.1		0.0653	0.0575	0.0140	27		033.01	5
034	Selenium, Total (Se) (ppm)	0563	0.1814	0.0015	0.2400	0.0596	0.0141	9	-0.98	034.04	0
034	Selenium, Total (Se) (ppm)	0045	0.1895	0.0270	0.2400	0.0596	0.0141	9	-0.85	034.04	0
034	Selenium, Total (Se) (ppm)	0171	0.2015	0.0030	0.2400	0.0596	0.0141	9	-0.65	034.04	0
034	Selenium, Total (Se) (ppm)	0164	0.2215	0.0010	0.2400	0.0596	0.0141	9	-0.31	034.53	0
034	Selenium, Total (Se) (ppm)	0169	0.2300	0.0200	0.2400	0.0596	0.0141	9	-0.17	034.04	0
034	Selenium, Total (Se) (ppm)	0553	0.2430	0.0180	0.2400	0.0596	0.0141	9	0.05	034.53	0
034	Selenium, Total (Se) (ppm)	0560	0.2450	0.0100	0.2400	0.0596	0.0141	9	0.08	034.53	0
034	Selenium, Total (Se) (ppm)	0407	0.3186	0.0407	0.2400	0.0596	0.0141	9	1.32	034.53	0
034	Selenium, Total (Se) (ppm)	0186	0.7370	0.0060	0.2400	0.0596	0.0141	9	8.34	034.52	0
034	Selenium, Total (Se) (ppm)	0692	< 1		0.2400	0.0596	0.0141	9		034.42	5
034	Selenium, Total (Se) (ppm)	0682	< 5		0.2400	0.0596	0.0141	9		034.43	5
035	Sodium (%)	0553	0.0033	0.0001	0.0169	0.0140	0.0060	44	-0.98	035.53	0
035	Sodium (%)	0154	0.0035	0.0001	0.0169	0.0140	0.0060	44	-0.96	035.52	0
035	Sodium (%)	0345	0.0035	0.0010	0.0169	0.0140	0.0060	44	-0.96	035.43	0
035	Sodium (%)	2326	0.0040	0.0000	0.0169	0.0140	0.0060	44	-0.92	035.43	0
035	Sodium (%)	0572	0.0041	0.0003	0.0169	0.0140	0.0060	44	-0.92	035.53	0
035	Sodium (%)	0407	0.0044	0.0002	0.0169	0.0140	0.0060	44	-0.90	035.53	0
035	Sodium (%)	0726	0.0050	0.0000	0.0169	0.0140	0.0060	44	-0.85	035.42	0
035	Sodium (%)	0083	0.0052	0.0003	0.0169	0.0140	0.0060	44	-0.84	035.43	0
035	Sodium (%)	0164	0.0055	0.0010	0.0169	0.0140	0.0060	44	-0.82	035.41	0
035	Sodium (%)	0870	0.0068	0.0002	0.0169	0.0140	0.0060	44	-0.72	035.42	0
035	Sodium (%)	0148	0.0070	0.0000	0.0169	0.0140	0.0060	44	-0.71	035.41	0
035	Sodium (%)	0870	0.0074	0.0003	0.0169	0.0140	0.0060	44	-0.68	035.43	0
035	Sodium (%)	0520	0.0075	0.0021	0.0169	0.0140	0.0060	44	-0.68	035.41	0
035	Sodium (%)	0693	0.0075	0.0010	0.0169	0.0140	0.0060	44	-0.67	035.42	0
035	Sodium (%)	0037	0.0085	0.0010	0.0169	0.0140	0.0060	44	-0.60	035.42	0
035	Sodium (%)	0918	0.0085	0.0010	0.0169	0.0140	0.0060	44	-0.60	035.43	0
035	Sodium (%)	0407	0.0093	0.0053	0.0169	0.0140	0.0060	44	-0.55	035.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
035	Sodium (%)	0263	0.0095	0.0010	0.0169	0.0140	0.0060	44	-0.53	035.42	0
035	Sodium (%)	0049	0.0100	0.0000	0.0169	0.0140	0.0060	44	-0.50	035.41	0
035	Sodium (%)	0226	0.0100	0.0000	0.0169	0.0140	0.0060	44	-0.50	035.41	0
035	Sodium (%)	0265	0.0100	0.0000	0.0169	0.0140	0.0060	44	-0.50	035.42	0
035	Sodium (%)	2089	0.0100	0.0000	0.0169	0.0140	0.0060	44	-0.50	035.43	0
035	Sodium (%)	0407	0.0115	0.0030	0.0169	0.0140	0.0060	44	-0.39	035.41	0
035	Sodium (%)	2259	0.0125	0.0030	0.0169	0.0140	0.0060	44	-0.32	035.41	0
035	Sodium (%)	0300	0.0141	0.0001	0.0169	0.0140	0.0060	44	-0.21	035.43	0
035	Sodium (%)	0366	0.0145	0.0050	0.0169	0.0140	0.0060	44	-0.17	035.42	0
035	Sodium (%)	0202	0.0150	0.0100	0.0169	0.0140	0.0060	44	-0.14	035.43	0
035	Sodium (%)	0144	0.0200	0.0000	0.0169	0.0140	0.0060	44	0.22	035.41	0
035	Sodium (%)	0692	0.0200	0.0000	0.0169	0.0140	0.0060	44	0.22	035.42	0
035	Sodium (%)	2097	0.0242	0.0029	0.0169	0.0140	0.0060	44	0.52	035.31	0
035	Sodium (%)	0536	0.0250	0.0100	0.0169	0.0140	0.0060	44	0.58	035.31	0
035	Sodium (%)	0171	0.0250	0.0100	0.0169	0.0140	0.0060	44	0.58	035.41	0
035	Sodium (%)	0229	0.0250	0.0100	0.0169	0.0140	0.0060	44	0.58	035.41	0
035	Sodium (%)	0123	0.0300	0.0000	0.0169	0.0140	0.0060	44	0.94	035.41	0
035	Sodium (%)	0425	0.0300	0.0000	0.0169	0.0140	0.0060	44	0.94	035.43	0
035	Sodium (%)	0682	0.0300	0.0000	0.0169	0.0140	0.0060	44	0.94	035.43	0
035	Sodium (%)	0142	0.0325	0.0019	0.0169	0.0140	0.0060	44	1.11	035.31	0
035	Sodium (%)	0278	0.0400	0.0000	0.0169	0.0140	0.0060	44	1.65	035.42	0
035	Sodium (%)	2188	0.0500	0.0000	0.0169	0.0140	0.0060	44	2.37	035.01	0
035	Sodium (%)	0353	0.0518	0.0145	0.0169	0.0140	0.0060	44	2.49	035.43	0
035	Sodium (%)	0186	0.0545	0.0871	0.0169	0.0140	0.0060	44	2.69	035.52	0
035	Sodium (%)	0598	0.0564	0.0892	0.0169	0.0140	0.0060	44	2.82	035.41	0
035	Sodium (%)	0510	0.0650	0.0020	0.0169	0.0140	0.0060	44	3.44	035.43	0
035	Sodium (%)	0169	0.2500	0.0000	0.0169	0.0140	0.0060	44	16.68	035.32	0
035	Sodium (%)	0042	0.0835	0.1330	0.0169	0.0140	0.0060	44	4.76	035.42	1
035	Sodium (%)	0354	0.3500	0.1000	0.0169	0.0140	0.0060	44	23.83	035.41	1
035	Sodium (%)	2129	1.855	0.0700	0.0169	0.0140	0.0060	44	131.52	035.42	2
035	Sodium (%)	0297	0.0000	0.0000	0.0169	0.0140	0.0060	44		035.43	4
035	Sodium (%)	0100	< 0.01		0.0169	0.0140	0.0060	44		035.99	5
036	Sulfur (%)	0366	0.3400	0.0200	0.3847	0.0259	0.0099	40	-1.73	036.42	0
036	Sulfur (%)	0242	0.3550	0.0100	0.3847	0.0259	0.0099	40	-1.15	036.99	0
036	Sulfur (%)	0045	0.3595	0.0230	0.3847	0.0259	0.0099	40	-0.98	036.42	0
036	Sulfur (%)	0278	0.3600	0.0200	0.3847	0.0259	0.0099	40	-0.96	036.42	0
036	Sulfur (%)	0692	0.3600	0.0200	0.3847	0.0259	0.0099	40	-0.96	036.42	0
036	Sulfur (%)	0693	0.3610	0.0180	0.3847	0.0259	0.0099	40	-0.92	036.42	0
036	Sulfur (%)	0870	0.3650	0.0213	0.3847	0.0259	0.0099	40	-0.76	036.42	0
036	Sulfur (%)	0035	0.3650	0.0100	0.3847	0.0259	0.0099	40	-0.76	036.42	0
036	Sulfur (%)	0353	0.3650	0.0300	0.3847	0.0259	0.0099	40	-0.76	036.43	0
036	Sulfur (%)	0357	0.3668	0.0027	0.3847	0.0259	0.0099	40	-0.70	036.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
036	Sulfur (%)	0870	0.3670	0.0250	0.3847	0.0259	0.0099	40	-0.69	036.43	0
036	Sulfur (%)	0186	0.3695	0.0130	0.3847	0.0259	0.0099	40	-0.59	036.42	0
036	Sulfur (%)	0553	0.3705	0.0090	0.3847	0.0259	0.0099	40	-0.55	036.53	0
036	Sulfur (%)	0148	0.3715	0.0090	0.3847	0.0259	0.0099	40	-0.51	036.04	0
036	Sulfur (%)	0407	0.3739	0.0417	0.3847	0.0259	0.0099	40	-0.42	036.42	0
036	Sulfur (%)	0226	0.3750	0.0100	0.3847	0.0259	0.0099	40	-0.38	036.04	0
036	Sulfur (%)	0229	0.3750	0.0100	0.3847	0.0259	0.0099	40	-0.38	036.04	0
036	Sulfur (%)	0164	0.3750	0.0100	0.3847	0.0259	0.0099	40	-0.38	036.42	0
036	Sulfur (%)	0186	0.3750	0.0158	0.3847	0.0259	0.0099	40	-0.38	036.52	0
036	Sulfur (%)	0726	0.3800	0.0000	0.3847	0.0259	0.0099	40	-0.18	036.42	0
036	Sulfur (%)	0083	0.3800	0.0000	0.3847	0.0259	0.0099	40	-0.18	036.43	0
036	Sulfur (%)	0297	0.3800	0.0000	0.3847	0.0259	0.0099	40	-0.18	036.43	0
036	Sulfur (%)	0918	0.3810	0.0020	0.3847	0.0259	0.0099	40	-0.14	036.43	0
036	Sulfur (%)	0407	0.3840	0.0043	0.3847	0.0259	0.0099	40	-0.03	036.43	0
036	Sulfur (%)	0265	0.3850	0.0100	0.3847	0.0259	0.0099	40	0.01	036.42	0
036	Sulfur (%)	0035	0.3850	0.0100	0.3847	0.0259	0.0099	40	0.01	036.43	0
036	Sulfur (%)	0202	0.3900	0.0000	0.3847	0.0259	0.0099	40	0.20	036.43	0
036	Sulfur (%)	0345	0.3900	0.0040	0.3847	0.0259	0.0099	40	0.20	036.43	0
036	Sulfur (%)	0098	0.3938	0.0014	0.3847	0.0259	0.0099	40	0.35	036.43	0
036	Sulfur (%)	0682	0.4000	0.0000	0.3847	0.0259	0.0099	40	0.59	036.43	0
036	Sulfur (%)	2326	0.4000	0.0000	0.3847	0.0259	0.0099	40	0.59	036.43	0
036	Sulfur (%)	0708	0.4035	0.0090	0.3847	0.0259	0.0099	40	0.73	036.42	0
036	Sulfur (%)	0169	0.4200	0.0200	0.3847	0.0259	0.0099	40	1.36	036.43	0
036	Sulfur (%)	2089	0.4200	0.0000	0.3847	0.0259	0.0099	40	1.36	036.43	0
036	Sulfur (%)	0964	0.4247	0.0053	0.3847	0.0259	0.0099	40	1.54	036.43	0
036	Sulfur (%)	0510	0.4300	0.0000	0.3847	0.0259	0.0099	40	1.75	036.43	0
036	Sulfur (%)	0037	0.4360	0.0020	0.3847	0.0259	0.0099	40	1.98	036.42	0
036	Sulfur (%)	0042	0.4400	0.0080	0.3847	0.0259	0.0099	40	2.14	036.42	0
036	Sulfur (%)	2129	0.4445	0.0030	0.3847	0.0259	0.0099	40	2.31	036.42	0
036	Sulfur (%)	0171	0.4500	0.0000	0.3847	0.0259	0.0099	40	2.53	036.42	0
037	Zinc (ppm)	0536	26.03	1.430	48.86	5.129	1.651	73	-4.45	037.31	0
037	Zinc (ppm)	0123	36.90	8.400	48.86	5.129	1.651	73	-2.33	037.41	0
037	Zinc (ppm)	0968	41.45	0.6270	48.86	5.129	1.651	73	-1.44	037.43	0
037	Zinc (ppm)	0066	41.45	0.6500	48.86	5.129	1.651	73	-1.44	037.44	0
037	Zinc (ppm)	0009	41.50	1.000	48.86	5.129	1.651	73	-1.43	037.42	0
037	Zinc (ppm)	0511	41.50	3.000	48.86	5.129	1.651	73	-1.43	037.43	0
037	Zinc (ppm)	0870	42.35	0.8600	48.86	5.129	1.651	73	-1.27	037.42	0
037	Zinc (ppm)	0202	42.50	1.000	48.86	5.129	1.651	73	-1.24	037.43	0
037	Zinc (ppm)	0598	42.65	0.6197	48.86	5.129	1.651	73	-1.21	037.41	0
037	Zinc (ppm)	0297	43.00	0.0000	48.86	5.129	1.651	73	-1.14	037.43	0
037	Zinc (ppm)	0870	43.01	0.4200	48.86	5.129	1.651	73	-1.14	037.43	0
037	Zinc (ppm)	0910	44.00	0.0000	48.86	5.129	1.651	73	-0.95	037.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
037	Zinc (ppm)	0242	44.00	0.0000	48.86	5.129	1.651	73	-0.95	037.99	0
037	Zinc (ppm)	0045	44.35	1.100	48.86	5.129	1.651	73	-0.88	037.42	0
037	Zinc (ppm)	2319	44.45	0.7000	48.86	5.129	1.651	73	-0.86	037.42	0
037	Zinc (ppm)	0164	44.50	1.000	48.86	5.129	1.651	73	-0.85	037.44	0
037	Zinc (ppm)	0263	45.10	0.6070	48.86	5.129	1.651	73	-0.73	037.42	0
037	Zinc (ppm)	0208	45.10	1.580	48.86	5.129	1.651	73	-0.73	037.41	0
037	Zinc (ppm)	0918	45.14	0.0100	48.86	5.129	1.651	73	-0.73	037.43	0
037	Zinc (ppm)	0848	45.71	0.3200	48.86	5.129	1.651	73	-0.61	037.41	0
037	Zinc (ppm)	0278	45.95	0.7000	48.86	5.129	1.651	73	-0.57	037.42	0
037	Zinc (ppm)	0083	46.00	0.0000	48.86	5.129	1.651	73	-0.56	037.43	0
037	Zinc (ppm)	0019	46.86	0.9400	48.86	5.129	1.651	73	-0.39	037.41	0
037	Zinc (ppm)	0229	46.99	1.860	48.86	5.129	1.651	73	-0.36	037.41	0
037	Zinc (ppm)	0100	47.00	0.0000	48.86	5.129	1.651	73	-0.36	037.99	0
037	Zinc (ppm)	0425	47.25	0.6900	48.86	5.129	1.651	73	-0.31	037.43	0
037	Zinc (ppm)	0653	47.48	0.2700	48.86	5.129	1.651	73	-0.27	037.41	0
037	Zinc (ppm)	0520	47.56	2.212	48.86	5.129	1.651	73	-0.25	037.41	0
037	Zinc (ppm)	0693	47.87	2.439	48.86	5.129	1.651	73	-0.19	037.42	0
037	Zinc (ppm)	0407	47.94	0.5729	48.86	5.129	1.651	73	-0.18	037.53	0
037	Zinc (ppm)	0037	48.00	1.200	48.86	5.129	1.651	73	-0.17	037.42	0
037	Zinc (ppm)	0013	48.05	0.9000	48.86	5.129	1.651	73	-0.16	037.43	0
037	Zinc (ppm)	0939	48.07	2.970	48.86	5.129	1.651	73	-0.15	037.53	0
037	Zinc (ppm)	0529	48.15	0.7000	48.86	5.129	1.651	73	-0.14	037.31	0
037	Zinc (ppm)	0345	48.55	0.1000	48.86	5.129	1.651	73	-0.06	037.43	0
037	Zinc (ppm)	0357	48.92	0.2040	48.86	5.129	1.651	73	0.01	037.42	0
037	Zinc (ppm)	0674	48.94	0.0000	48.86	5.129	1.651	73	0.02	037.41	0
037	Zinc (ppm)	0407	48.98	1.387	48.86	5.129	1.651	73	0.02	037.43	0
037	Zinc (ppm)	0049	49.06	6.880	48.86	5.129	1.651	73	0.04	037.41	0
037	Zinc (ppm)	0354	49.30	2.590	48.86	5.129	1.651	73	0.09	037.41	0
037	Zinc (ppm)	0208	49.35	0.1000	48.86	5.129	1.651	73	0.10	037.31	0
037	Zinc (ppm)	0553	49.45	0.3000	48.86	5.129	1.651	73	0.12	037.53	0
037	Zinc (ppm)	2326	49.50	3.000	48.86	5.129	1.651	73	0.13	037.43	0
037	Zinc (ppm)	0098	49.73	0.1000	48.86	5.129	1.651	73	0.17	037.41	0
037	Zinc (ppm)	0265	49.80	7.400	48.86	5.129	1.651	73	0.18	037.42	0
037	Zinc (ppm)	0010	50.00	0.0000	48.86	5.129	1.651	73	0.22	037.33	0
037	Zinc (ppm)	0692	50.50	4.800	48.86	5.129	1.651	73	0.32	037.42	0
037	Zinc (ppm)	2089	50.52	0.8700	48.86	5.129	1.651	73	0.32	037.43	0
037	Zinc (ppm)	0169	50.85	1.500	48.86	5.129	1.651	73	0.39	037.43	0
037	Zinc (ppm)	0407	50.90	0.8732	48.86	5.129	1.651	73	0.40	037.41	0
037	Zinc (ppm)	0510	51.00	0.0000	48.86	5.129	1.651	73	0.42	037.43	0
037	Zinc (ppm)	2188	51.22	0.0000	48.86	5.129	1.651	73	0.46	037.31	0
037	Zinc (ppm)	0726	51.40	0.0200	48.86	5.129	1.651	73	0.50	037.42	0
037	Zinc (ppm)	0035	51.50	1.000	48.86	5.129	1.651	73	0.52	037.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
037	Zinc (ppm)	0033	51.61	0.0100	48.86	5.129	1.651	73	0.54	037.43	0
037	Zinc (ppm)	0226	51.62	1.490	48.86	5.129	1.651	73	0.54	037.41	0
037	Zinc (ppm)	0190	53.03	1.840	48.86	5.129	1.651	73	0.81	037.42	0
037	Zinc (ppm)	0004	53.14	2.320	48.86	5.129	1.651	73	0.84	037.41	0
037	Zinc (ppm)	0366	53.50	5.000	48.86	5.129	1.651	73	0.91	037.42	0
037	Zinc (ppm)	0171	53.51	1.020	48.86	5.129	1.651	73	0.91	037.41	0
037	Zinc (ppm)	0572	53.80	6.400	48.86	5.129	1.651	73	0.96	037.53	0
037	Zinc (ppm)	0651	53.95	1.710	48.86	5.129	1.651	73	0.99	037.31	0
037	Zinc (ppm)	0563	54.17	0.1528	48.86	5.129	1.651	73	1.04	037.31	0
037	Zinc (ppm)	0035	54.50	1.000	48.86	5.129	1.651	73	1.10	037.42	0
037	Zinc (ppm)	0148	55.41	1.710	48.86	5.129	1.651	73	1.28	037.41	0
037	Zinc (ppm)	0047	55.75	8.060	48.86	5.129	1.651	73	1.34	037.52	0
037	Zinc (ppm)	0505	56.73	0.8800	48.86	5.129	1.651	73	1.54	037.43	0
037	Zinc (ppm)	0011	59.80	1.400	48.86	5.129	1.651	73	2.13	037.43	0
037	Zinc (ppm)	0042	60.05	1.500	48.86	5.129	1.651	73	2.18	037.42	0
037	Zinc (ppm)	0027	60.93	8.142	48.86	5.129	1.651	73	2.35	037.43	0
037	Zinc (ppm)	2302	66.00	0.0000	48.86	5.129	1.651	73	3.34	037.99	0
037	Zinc (ppm)	0682	71.56	0.0000	48.86	5.129	1.651	73	4.43	037.43	0
037	Zinc (ppm)	2097	178.0	8.000	48.86	5.129	1.651	73	25.18	037.31	0
037	Zinc (ppm)	0353	92.00	12.00	48.86	5.129	1.651	73	8.41	037.43	1
037	Zinc (ppm)	2129	1,090	16.00	48.86	5.129	1.651	73	202.99	037.42	2
038	Molybdenum (ppm)	0083	2.000	0.0000	2.538	0.3821	0.1168	18	-1.41	038.43	0
038	Molybdenum (ppm)	0297	2.110	0.2600	2.538	0.3821	0.1168	18	-1.12	038.43	0
038	Molybdenum (ppm)	0407	2.137	0.0162	2.538	0.3821	0.1168	18	-1.05	038.41	0
038	Molybdenum (ppm)	0171	2.176	0.0130	2.538	0.3821	0.1168	18	-0.95	038.41	0
038	Molybdenum (ppm)	0345	2.185	0.0300	2.538	0.3821	0.1168	18	-0.92	038.43	0
038	Molybdenum (ppm)	0407	2.314	0.2205	2.538	0.3821	0.1168	18	-0.58	038.53	0
038	Molybdenum (ppm)	0407	2.460	0.0701	2.538	0.3821	0.1168	18	-0.20	038.43	0
038	Molybdenum (ppm)	0278	2.495	0.0300	2.538	0.3821	0.1168	18	-0.11	038.42	0
038	Molybdenum (ppm)	0553	2.525	0.0300	2.538	0.3821	0.1168	18	-0.03	038.53	0
038	Molybdenum (ppm)	0202	2.605	0.1300	2.538	0.3821	0.1168	18	0.18	038.43	0
038	Molybdenum (ppm)	0164	2.610	0.0400	2.538	0.3821	0.1168	18	0.19	038.53	0
038	Molybdenum (ppm)	0693	2.613	0.1430	2.538	0.3821	0.1168	18	0.20	038.42	0
038	Molybdenum (ppm)	0226	2.675	0.0300	2.538	0.3821	0.1168	18	0.36	038.41	0
038	Molybdenum (ppm)	0964	2.785	0.0900	2.538	0.3821	0.1168	18	0.65	038.43	0
038	Molybdenum (ppm)	0510	2.850	0.1000	2.538	0.3821	0.1168	18	0.82	038.43	0
038	Molybdenum (ppm)	0169	2.935	0.0100	2.538	0.3821	0.1168	18	1.04	038.43	0
038	Molybdenum (ppm)	0045	3.090	0.1200	2.538	0.3821	0.1168	18	1.45	038.42	0
038	Molybdenum (ppm)	0265	4.575	0.7700	2.538	0.3821	0.1168	18	5.33	038.42	0
038	Molybdenum (ppm)	0123	44.50	1.400	2.538	0.3821	0.1168	18	109.82	038.41	2
040	Barium (ppm)	0407	7.062	0.7513				1		040.53	0
041	Vanadium (ppm)	0553	0.0309	0.0003				2		041.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
041	Vanadium (ppm)	0407	0.0446	0.0045				2		041.53	0
042	Chloride (%)	0948	0.0200	0.0000				3		042.00	0
042	Chloride (%)	0226	0.0300	0.0000				3		042.00	0
042	Chloride (%)	2188	0.0300	0.0000				3		042.01	0
042	Chloride (%)	0297	0.0000	0.0000				3		042.99	4
101	Choline Chloride (ppm)	0227	3.000	0.2000				1		101.99	0
102	Niacin (ppm)	0227	25.95	0.7000				1		102.01	0
103	Pantothenic Acid (ppm)	0227	10.90	0.4000				1		103.01	0
104	Riboflavin (ppm)	0227	2.780	0.0400				1		104.00	0
105	Thiamine (ppm)	0227	3.165	0.0100				1		105.01	0
106	Vitamin A (KU / kg)	0227	< 0.3					0		106.00	5
106	Vitamin A (KU / kg)	0098	< 0.8					0		106.01	5
107	Vitamin B12 (ppb)	0227	< 4.4					0		107.00	5
108	Vitamin D3 (KU / kg)	0227	< 0.04					0		108.02	5
109	Vitamin E (IU / kg)	0227	12.70	0.2000				2		109.02	0
109	Vitamin E (IU / kg)	0563	14.32	0.5783				2		109.02	0
109	Vitamin E (IU / kg)	0098	< 40					2		109.02	5
111	Vitamin C, Phosphorylated (ppm)	0227	< 4.4					0		111.00	5
112	Pyridoxine (µg / g)	0227	5.690	0.0600				1		112.01	0
113	Folic Acid (ppm)	0227	2.815	0.2300				1		113.01	0
114	Biotin (ppm)	0227	0.4350	0.0180				1		114.01	0
115	Non Protein N (NPN) (%)	0098	1.005	0.0300				1		115.00	0
118	Peroxide value (meq/kg)	0853	3.800	1.200				1		118.99	0
120	Alanine (%)	0353	1.790	0.0200	2.016	0.0780	0.0248	25	-2.90	120.00	0
120	Alanine (%)	2246	1.840	0.0000	2.016	0.0780	0.0248	25	-2.26	120.05	0
120	Alanine (%)	0407	1.859	0.0630	2.016	0.0780	0.0248	25	-2.02	120.05	0
120	Alanine (%)	0242	1.860	0.1400	2.016	0.0780	0.0248	25	-2.00	120.05	0
120	Alanine (%)	0918	1.942	0.0100	2.016	0.0780	0.0248	25	-0.95	120.00	0
120	Alanine (%)	0676	1.945	0.0100	2.016	0.0780	0.0248	25	-0.91	120.05	0
120	Alanine (%)	0968	1.975	0.0100	2.016	0.0780	0.0248	25	-0.52	120.00	0
120	Alanine (%)	0652	2.000	0.0000	2.016	0.0780	0.0248	25	-0.20	120.00	0
120	Alanine (%)	0941	2.004	0.0000	2.016	0.0780	0.0248	25	-0.15	120.00	0
120	Alanine (%)	0504	2.010	0.0200	2.016	0.0780	0.0248	25	-0.08	120.00	0
120	Alanine (%)	0682	2.012	0.0000	2.016	0.0780	0.0248	25	-0.05	120.00	0
120	Alanine (%)	0226	2.020	0.0800	2.016	0.0780	0.0248	25	0.05	120.00	0
120	Alanine (%)	0684	2.025	0.0100	2.016	0.0780	0.0248	25	0.12	120.00	0
120	Alanine (%)	0910	2.030	0.0600	2.016	0.0780	0.0248	25	0.18	120.00	0
120	Alanine (%)	2319	2.035	0.0100	2.016	0.0780	0.0248	25	0.25	120.05	0
120	Alanine (%)	2059	2.044	0.0020	2.016	0.0780	0.0248	25	0.36	120.00	0
120	Alanine (%)	2188	2.054	0.0000	2.016	0.0780	0.0248	25	0.49	120.05	0
120	Alanine (%)	0675	2.070	0.0200	2.016	0.0780	0.0248	25	0.69	120.00	0
120	Alanine (%)	0571	2.078	0.0380	2.016	0.0780	0.0248	25	0.80	120.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
120	Alanine (%)	0644	2.081	0.0070	2.016	0.0780	0.0248	25	0.83	120.00	0
120	Alanine (%)	0354	2.081	0.0020	2.016	0.0780	0.0248	25	0.84	120.00	0
120	Alanine (%)	0098	2.086	0.0280	2.016	0.0780	0.0248	25	0.90	120.02	0
120	Alanine (%)	0723	2.095	0.0300	2.016	0.0780	0.0248	25	1.01	120.00	0
120	Alanine (%)	0227	2.100	0.0000	2.016	0.0780	0.0248	25	1.08	120.99	0
120	Alanine (%)	0859	2.115	0.0589	2.016	0.0780	0.0248	25	1.27	120.00	0
120	Alanine (%)	0148	2.005	0.2100	2.016	0.0780	0.0248	25	-0.14	120.05	1
121	Arginine (%)	0353	3.100	0.0000	3.354	0.0989	0.0400	25	-2.57	121.00	0
121	Arginine (%)	2246	3.170	0.1000	3.354	0.0989	0.0400	25	-1.86	121.05	0
121	Arginine (%)	0226	3.195	0.0300	3.354	0.0989	0.0400	25	-1.61	121.00	0
121	Arginine (%)	0918	3.196	0.0750	3.354	0.0989	0.0400	25	-1.61	121.00	0
121	Arginine (%)	0684	3.285	0.1100	3.354	0.0989	0.0400	25	-0.70	121.00	0
121	Arginine (%)	0098	3.299	0.0280	3.354	0.0989	0.0400	25	-0.56	121.02	0
121	Arginine (%)	0407	3.307	0.0390	3.354	0.0989	0.0400	25	-0.48	121.05	0
121	Arginine (%)	0652	3.325	0.0100	3.354	0.0989	0.0400	25	-0.30	121.00	0
121	Arginine (%)	0675	3.330	0.0400	3.354	0.0989	0.0400	25	-0.25	121.00	0
121	Arginine (%)	0968	3.330	0.0400	3.354	0.0989	0.0400	25	-0.25	121.00	0
121	Arginine (%)	0504	3.345	0.0100	3.354	0.0989	0.0400	25	-0.09	121.00	0
121	Arginine (%)	0941	3.350	0.0700	3.354	0.0989	0.0400	25	-0.04	121.00	0
121	Arginine (%)	0676	3.350	0.0000	3.354	0.0989	0.0400	25	-0.04	121.05	0
121	Arginine (%)	0910	3.365	0.0900	3.354	0.0989	0.0400	25	0.11	121.00	0
121	Arginine (%)	0354	3.379	0.0020	3.354	0.0989	0.0400	25	0.25	121.00	0
121	Arginine (%)	2059	3.384	0.0020	3.354	0.0989	0.0400	25	0.30	121.00	0
121	Arginine (%)	0682	3.392	0.0000	3.354	0.0989	0.0400	25	0.38	121.00	0
121	Arginine (%)	0227	3.415	0.0300	3.354	0.0989	0.0400	25	0.61	121.99	0
121	Arginine (%)	0859	3.419	0.0824	3.354	0.0989	0.0400	25	0.66	121.00	0
121	Arginine (%)	2319	3.420	0.0000	3.354	0.0989	0.0400	25	0.66	121.05	0
121	Arginine (%)	0723	3.440	0.0800	3.354	0.0989	0.0400	25	0.87	121.00	0
121	Arginine (%)	2188	3.448	0.0000	3.354	0.0989	0.0400	25	0.95	121.05	0
121	Arginine (%)	0644	3.463	0.0030	3.354	0.0989	0.0400	25	1.09	121.00	0
121	Arginine (%)	0242	3.485	0.1300	3.354	0.0989	0.0400	25	1.32	121.05	0
121	Arginine (%)	0571	3.533	0.0290	3.354	0.0989	0.0400	25	1.80	121.00	0
121	Arginine (%)	0148	3.265	0.3100	3.354	0.0989	0.0400	25	-0.90	121.05	1
122	Aspartic (%)	0407	4.454	0.2180	5.354	0.2257	0.0525	24	-3.99	122.05	0
122	Aspartic (%)	2246	4.715	0.0500	5.354	0.2257	0.0525	24	-2.83	122.05	0
122	Aspartic (%)	2188	4.905	0.0000	5.354	0.2257	0.0525	24	-1.99	122.05	0
122	Aspartic (%)	0941	5.085	0.0500	5.354	0.2257	0.0525	24	-1.19	122.00	0
122	Aspartic (%)	0918	5.167	0.0190	5.354	0.2257	0.0525	24	-0.83	122.00	0
122	Aspartic (%)	0652	5.255	0.0300	5.354	0.2257	0.0525	24	-0.44	122.00	0
122	Aspartic (%)	0682	5.268	0.0000	5.354	0.2257	0.0525	24	-0.38	122.00	0
122	Aspartic (%)	0684	5.285	0.0300	5.354	0.2257	0.0525	24	-0.30	122.00	0
122	Aspartic (%)	0504	5.300	0.0200	5.354	0.2257	0.0525	24	-0.24	122.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
122	Aspartic (%)	0354	5.317	0.0100	5.354	0.2257	0.0525	24	-0.16	122.00	0
122	Aspartic (%)	0675	5.320	0.0200	5.354	0.2257	0.0525	24	-0.15	122.00	0
122	Aspartic (%)	0910	5.330	0.0200	5.354	0.2257	0.0525	24	-0.10	122.00	0
122	Aspartic (%)	0098	5.358	0.0340	5.354	0.2257	0.0525	24	0.02	122.02	0
122	Aspartic (%)	2059	5.379	0.0200	5.354	0.2257	0.0525	24	0.11	122.00	0
122	Aspartic (%)	0859	5.417	0.1230	5.354	0.2257	0.0525	24	0.28	122.00	0
122	Aspartic (%)	0968	5.430	0.0200	5.354	0.2257	0.0525	24	0.34	122.00	0
122	Aspartic (%)	0571	5.469	0.0030	5.354	0.2257	0.0525	24	0.51	122.00	0
122	Aspartic (%)	0644	5.477	0.0030	5.354	0.2257	0.0525	24	0.54	122.00	0
122	Aspartic (%)	0353	5.510	0.0200	5.354	0.2257	0.0525	24	0.69	122.00	0
122	Aspartic (%)	0723	5.520	0.0800	5.354	0.2257	0.0525	24	0.74	122.00	0
122	Aspartic (%)	0227	5.535	0.0300	5.354	0.2257	0.0525	24	0.80	122.99	0
122	Aspartic (%)	2319	5.660	0.0800	5.354	0.2257	0.0525	24	1.36	122.05	0
122	Aspartic (%)	0226	5.670	0.2400	5.354	0.2257	0.0525	24	1.40	122.00	0
122	Aspartic (%)	0676	5.720	0.1400	5.354	0.2257	0.0525	24	1.62	122.05	0
122	Aspartic (%)	0242	4.800	0.4600	5.354	0.2257	0.0525	24	-2.45	122.05	1
122	Aspartic (%)	0148	5.415	0.4700	5.354	0.2257	0.0525	24	0.27	122.05	1
124	Cysteine/Cystine (%)	0353	0.5600	0.0200	0.6545	0.0545	0.0159	25	-1.73	124.00	0
124	Cysteine/Cystine (%)	2246	0.5800	0.0000	0.6545	0.0545	0.0159	25	-1.37	124.05	0
124	Cysteine/Cystine (%)	0652	0.5900	0.0000	0.6545	0.0545	0.0159	25	-1.18	124.00	0
124	Cysteine/Cystine (%)	2059	0.6080	0.0060	0.6545	0.0545	0.0159	25	-0.85	124.00	0
124	Cysteine/Cystine (%)	2319	0.6100	0.0000	0.6545	0.0545	0.0159	25	-0.82	124.05	0
124	Cysteine/Cystine (%)	0407	0.6194	0.0787	0.6545	0.0545	0.0159	25	-0.64	124.05	0
124	Cysteine/Cystine (%)	0941	0.6235	0.0130	0.6545	0.0545	0.0159	25	-0.57	124.00	0
124	Cysteine/Cystine (%)	0684	0.6250	0.0100	0.6545	0.0545	0.0159	25	-0.54	124.00	0
124	Cysteine/Cystine (%)	0227	0.6300	0.0000	0.6545	0.0545	0.0159	25	-0.45	124.99	0
124	Cysteine/Cystine (%)	0644	0.6305	0.0030	0.6545	0.0545	0.0159	25	-0.44	124.00	0
124	Cysteine/Cystine (%)	0859	0.6363	0.0040	0.6545	0.0545	0.0159	25	-0.33	124.00	0
124	Cysteine/Cystine (%)	2188	0.6390	0.0000	0.6545	0.0545	0.0159	25	-0.28	124.05	0
124	Cysteine/Cystine (%)	0910	0.6450	0.0100	0.6545	0.0545	0.0159	25	-0.17	124.00	0
124	Cysteine/Cystine (%)	0918	0.6450	0.0300	0.6545	0.0545	0.0159	25	-0.17	124.00	0
124	Cysteine/Cystine (%)	0354	0.6510	0.0080	0.6545	0.0545	0.0159	25	-0.06	124.00	0
124	Cysteine/Cystine (%)	0098	0.6735	0.0170	0.6545	0.0545	0.0159	25	0.35	124.02	0
124	Cysteine/Cystine (%)	0571	0.6855	0.0170	0.6545	0.0545	0.0159	25	0.57	124.00	0
124	Cysteine/Cystine (%)	0504	0.6950	0.0100	0.6545	0.0545	0.0159	25	0.74	124.00	0
124	Cysteine/Cystine (%)	0968	0.6950	0.0100	0.6545	0.0545	0.0159	25	0.74	124.00	0
124	Cysteine/Cystine (%)	0226	0.7000	0.0200	0.6545	0.0545	0.0159	25	0.83	124.00	0
124	Cysteine/Cystine (%)	0676	0.7050	0.0100	0.6545	0.0545	0.0159	25	0.93	124.05	0
124	Cysteine/Cystine (%)	0723	0.7100	0.0200	0.6545	0.0545	0.0159	25	1.02	124.00	0
124	Cysteine/Cystine (%)	0682	0.7200	0.0000	0.6545	0.0545	0.0159	25	1.20	124.00	0
124	Cysteine/Cystine (%)	0242	0.7650	0.0500	0.6545	0.0545	0.0159	25	2.03	124.05	0
124	Cysteine/Cystine (%)	0675	0.9300	0.0600	0.6545	0.0545	0.0159	25	5.05	124.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
124	Cysteine/Cystine (%)	0148	0.5175	0.1190	0.6545	0.0545	0.0159	25	-2.51	124.05	1
125	Glutamic (%)	0407	7.480	0.1780	8.589	0.2776	0.0925	24	-3.99	125.05	0
125	Glutamic (%)	2246	7.855	0.0300	8.589	0.2776	0.0925	24	-2.64	125.05	0
125	Glutamic (%)	0918	8.228	0.0290	8.589	0.2776	0.0925	24	-1.30	125.00	0
125	Glutamic (%)	0910	8.420	0.1800	8.589	0.2776	0.0925	24	-0.61	125.00	0
125	Glutamic (%)	2188	8.438	0.0000	8.589	0.2776	0.0925	24	-0.54	125.05	0
125	Glutamic (%)	0652	8.440	0.0000	8.589	0.2776	0.0925	24	-0.54	125.00	0
125	Glutamic (%)	0098	8.449	0.1130	8.589	0.2776	0.0925	24	-0.51	125.02	0
125	Glutamic (%)	0675	8.455	0.0700	8.589	0.2776	0.0925	24	-0.48	125.00	0
125	Glutamic (%)	2059	8.461	0.0350	8.589	0.2776	0.0925	24	-0.46	125.00	0
125	Glutamic (%)	0684	8.465	0.0500	8.589	0.2776	0.0925	24	-0.45	125.00	0
125	Glutamic (%)	0504	8.470	0.0600	8.589	0.2776	0.0925	24	-0.43	125.00	0
125	Glutamic (%)	0676	8.510	0.2600	8.589	0.2776	0.0925	24	-0.28	125.05	0
125	Glutamic (%)	0968	8.545	0.0300	8.589	0.2776	0.0925	24	-0.16	125.00	0
125	Glutamic (%)	0941	8.575	0.0000	8.589	0.2776	0.0925	24	-0.05	125.00	0
125	Glutamic (%)	0354	8.646	0.0460	8.589	0.2776	0.0925	24	0.21	125.00	0
125	Glutamic (%)	2319	8.730	0.0400	8.589	0.2776	0.0925	24	0.51	125.05	0
125	Glutamic (%)	0571	8.736	0.2090	8.589	0.2776	0.0925	24	0.53	125.00	0
125	Glutamic (%)	0353	8.750	0.0800	8.589	0.2776	0.0925	24	0.58	125.00	0
125	Glutamic (%)	0859	8.798	0.1928	8.589	0.2776	0.0925	24	0.75	125.00	0
125	Glutamic (%)	0644	8.800	0.0070	8.589	0.2776	0.0925	24	0.76	125.00	0
125	Glutamic (%)	0227	8.940	0.0400	8.589	0.2776	0.0925	24	1.27	125.99	0
125	Glutamic (%)	0723	8.955	0.1500	8.589	0.2776	0.0925	24	1.32	125.00	0
125	Glutamic (%)	0682	8.971	0.0000	8.589	0.2776	0.0925	24	1.38	125.00	0
125	Glutamic (%)	0226	9.070	0.4200	8.589	0.2776	0.0925	24	1.73	125.00	0
125	Glutamic (%)	0148	8.790	0.7400	8.589	0.2776	0.0925	24	0.73	125.05	1
126	Glycine (%)	0948	1.720	0.1440	1.966	0.0638	0.0336	26	-3.86	126.01	0
126	Glycine (%)	0918	1.850	0.0050	1.966	0.0638	0.0336	26	-1.83	126.00	0
126	Glycine (%)	2246	1.890	0.0000	1.966	0.0638	0.0336	26	-1.19	126.05	0
126	Glycine (%)	0676	1.895	0.0500	1.966	0.0638	0.0336	26	-1.11	126.05	0
126	Glycine (%)	0353	1.900	0.0400	1.966	0.0638	0.0336	26	-1.04	126.00	0
126	Glycine (%)	2319	1.900	0.0200	1.966	0.0638	0.0336	26	-1.04	126.05	0
126	Glycine (%)	0652	1.925	0.0300	1.966	0.0638	0.0336	26	-0.64	126.00	0
126	Glycine (%)	0684	1.940	0.0200	1.966	0.0638	0.0336	26	-0.41	126.00	0
126	Glycine (%)	2188	1.949	0.0000	1.966	0.0638	0.0336	26	-0.27	126.05	0
126	Glycine (%)	0910	1.950	0.0400	1.966	0.0638	0.0336	26	-0.25	126.00	0
126	Glycine (%)	0242	1.950	0.0800	1.966	0.0638	0.0336	26	-0.25	126.05	0
126	Glycine (%)	0968	1.955	0.0100	1.966	0.0638	0.0336	26	-0.17	126.00	0
126	Glycine (%)	0941	1.966	0.0600	1.966	0.0638	0.0336	26	0.00	126.00	0
126	Glycine (%)	0504	1.970	0.0200	1.966	0.0638	0.0336	26	0.06	126.00	0
126	Glycine (%)	2059	1.978	0.0080	1.966	0.0638	0.0336	26	0.19	126.00	0
126	Glycine (%)	0723	1.990	0.0200	1.966	0.0638	0.0336	26	0.38	126.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
126	Glycine (%)	0354	1.992	0.0170	1.966	0.0638	0.0336	26	0.40	126.00	0
126	Glycine (%)	0226	1.995	0.0700	1.966	0.0638	0.0336	26	0.45	126.00	0
126	Glycine (%)	0675	1.995	0.0100	1.966	0.0638	0.0336	26	0.45	126.00	0
126	Glycine (%)	0098	2.000	0.0840	1.966	0.0638	0.0336	26	0.53	126.02	0
126	Glycine (%)	0682	2.019	0.0000	1.966	0.0638	0.0336	26	0.83	126.00	0
126	Glycine (%)	0571	2.028	0.0450	1.966	0.0638	0.0336	26	0.96	126.00	0
126	Glycine (%)	0859	2.036	0.0596	1.966	0.0638	0.0336	26	1.09	126.00	0
126	Glycine (%)	0644	2.044	0.0080	1.966	0.0638	0.0336	26	1.22	126.00	0
126	Glycine (%)	0227	2.050	0.0200	1.966	0.0638	0.0336	26	1.32	126.99	0
126	Glycine (%)	0407	2.179	0.0140	1.966	0.0638	0.0336	26	3.34	126.05	0
126	Glycine (%)	0148	1.885	0.1900	1.966	0.0638	0.0336	26	-1.27	126.05	1
127	Histidine (%)	0676	1.005	0.0900	1.209	0.0646	0.0227	25	-3.15	127.05	0
127	Histidine (%)	0226	1.105	0.0500	1.209	0.0646	0.0227	25	-1.60	127.00	0
127	Histidine (%)	0918	1.116	0.0170	1.209	0.0646	0.0227	25	-1.44	127.00	0
127	Histidine (%)	2188	1.132	0.0000	1.209	0.0646	0.0227	25	-1.19	127.05	0
127	Histidine (%)	0682	1.133	0.0000	1.209	0.0646	0.0227	25	-1.17	127.00	0
127	Histidine (%)	0504	1.145	0.0100	1.209	0.0646	0.0227	25	-0.98	127.00	0
127	Histidine (%)	0910	1.180	0.0400	1.209	0.0646	0.0227	25	-0.44	127.00	0
127	Histidine (%)	0354	1.197	0.0060	1.209	0.0646	0.0227	25	-0.18	127.00	0
127	Histidine (%)	0684	1.200	0.0200	1.209	0.0646	0.0227	25	-0.13	127.00	0
127	Histidine (%)	0675	1.205	0.0100	1.209	0.0646	0.0227	25	-0.06	127.00	0
127	Histidine (%)	0098	1.206	0.0120	1.209	0.0646	0.0227	25	-0.04	127.02	0
127	Histidine (%)	0353	1.210	0.0200	1.209	0.0646	0.0227	25	0.02	127.00	0
127	Histidine (%)	0723	1.210	0.0000	1.209	0.0646	0.0227	25	0.02	127.00	0
127	Histidine (%)	2246	1.215	0.0500	1.209	0.0646	0.0227	25	0.10	127.05	0
127	Histidine (%)	2059	1.218	0.0060	1.209	0.0646	0.0227	25	0.15	127.00	0
127	Histidine (%)	0859	1.223	0.0307	1.209	0.0646	0.0227	25	0.23	127.00	0
127	Histidine (%)	0571	1.249	0.0040	1.209	0.0646	0.0227	25	0.63	127.00	0
127	Histidine (%)	0968	1.250	0.0200	1.209	0.0646	0.0227	25	0.64	127.00	0
127	Histidine (%)	0941	1.253	0.0380	1.209	0.0646	0.0227	25	0.69	127.00	0
127	Histidine (%)	0644	1.255	0.0010	1.209	0.0646	0.0227	25	0.71	127.00	0
127	Histidine (%)	0652	1.265	0.0500	1.209	0.0646	0.0227	25	0.87	127.00	0
127	Histidine (%)	2319	1.265	0.0100	1.209	0.0646	0.0227	25	0.87	127.05	0
127	Histidine (%)	0227	1.275	0.0100	1.209	0.0646	0.0227	25	1.03	127.99	0
127	Histidine (%)	0242	1.285	0.0500	1.209	0.0646	0.0227	25	1.18	127.05	0
127	Histidine (%)	0407	1.357	0.0220	1.209	0.0646	0.0227	25	2.30	127.05	0
127	Histidine (%)	0148	1.200	0.1600	1.209	0.0646	0.0227	25	-0.13	127.05	1
128	Isoleucine (%)	2319	1.690	0.0200	2.095	0.1401	0.0357	24	-2.89	128.05	0
128	Isoleucine (%)	0353	1.825	0.0300	2.095	0.1401	0.0357	24	-1.92	128.00	0
128	Isoleucine (%)	0918	1.909	0.0030	2.095	0.1401	0.0357	24	-1.33	128.00	0
128	Isoleucine (%)	0226	1.975	0.1500	2.095	0.1401	0.0357	24	-0.85	128.00	0
128	Isoleucine (%)	0723	1.975	0.1100	2.095	0.1401	0.0357	24	-0.85	128.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
128	Isoleucine (%)	0354	1.975	0.0040	2.095	0.1401	0.0357	24	-0.85	128.00	0
128	Isoleucine (%)	0676	1.985	0.0300	2.095	0.1401	0.0357	24	-0.78	128.05	0
128	Isoleucine (%)	2246	2.045	0.0100	2.095	0.1401	0.0357	24	-0.35	128.05	0
128	Isoleucine (%)	0652	2.055	0.0100	2.095	0.1401	0.0357	24	-0.28	128.00	0
128	Isoleucine (%)	0968	2.055	0.0100	2.095	0.1401	0.0357	24	-0.28	128.00	0
128	Isoleucine (%)	0910	2.090	0.0400	2.095	0.1401	0.0357	24	-0.03	128.00	0
128	Isoleucine (%)	0407	2.108	0.0680	2.095	0.1401	0.0357	24	0.09	128.05	0
128	Isoleucine (%)	0684	2.115	0.1100	2.095	0.1401	0.0357	24	0.14	128.00	0
128	Isoleucine (%)	0859	2.120	0.0330	2.095	0.1401	0.0357	24	0.18	128.00	0
128	Isoleucine (%)	0682	2.130	0.0000	2.095	0.1401	0.0357	24	0.25	128.00	0
128	Isoleucine (%)	2059	2.160	0.0020	2.095	0.1401	0.0357	24	0.47	128.00	0
128	Isoleucine (%)	0098	2.161	0.0390	2.095	0.1401	0.0357	24	0.47	128.02	0
128	Isoleucine (%)	0941	2.163	0.0510	2.095	0.1401	0.0357	24	0.48	128.00	0
128	Isoleucine (%)	2188	2.191	0.0000	2.095	0.1401	0.0357	24	0.69	128.05	0
128	Isoleucine (%)	0571	2.214	0.0510	2.095	0.1401	0.0357	24	0.85	128.00	0
128	Isoleucine (%)	0227	2.215	0.0300	2.095	0.1401	0.0357	24	0.86	128.99	0
128	Isoleucine (%)	0675	2.255	0.0100	2.095	0.1401	0.0357	24	1.14	128.00	0
128	Isoleucine (%)	0504	2.320	0.0400	2.095	0.1401	0.0357	24	1.61	128.00	0
128	Isoleucine (%)	0644	2.333	0.0060	2.095	0.1401	0.0357	24	1.70	128.00	0
128	Isoleucine (%)	0148	2.130	0.3000	2.095	0.1401	0.0357	24	0.25	128.05	1
129	Leucine (%)	2319	3.330	0.0000	3.560	0.1120	0.0390	25	-2.06	129.05	0
129	Leucine (%)	0407	3.348	0.1090	3.560	0.1120	0.0390	25	-1.90	129.05	0
129	Leucine (%)	0918	3.439	0.0400	3.560	0.1120	0.0390	25	-1.08	129.00	0
129	Leucine (%)	0676	3.445	0.0300	3.560	0.1120	0.0390	25	-1.03	129.05	0
129	Leucine (%)	0652	3.455	0.0100	3.560	0.1120	0.0390	25	-0.94	129.00	0
129	Leucine (%)	0684	3.460	0.0000	3.560	0.1120	0.0390	25	-0.90	129.00	0
129	Leucine (%)	0242	3.485	0.1700	3.560	0.1120	0.0390	25	-0.67	129.05	0
129	Leucine (%)	2246	3.495	0.0100	3.560	0.1120	0.0390	25	-0.58	129.05	0
129	Leucine (%)	0910	3.515	0.0500	3.560	0.1120	0.0390	25	-0.41	129.00	0
129	Leucine (%)	0968	3.530	0.0600	3.560	0.1120	0.0390	25	-0.27	129.00	0
129	Leucine (%)	0941	3.540	0.0660	3.560	0.1120	0.0390	25	-0.18	129.00	0
129	Leucine (%)	0723	3.550	0.0200	3.560	0.1120	0.0390	25	-0.09	129.00	0
129	Leucine (%)	2059	3.564	0.0040	3.560	0.1120	0.0390	25	0.03	129.00	0
129	Leucine (%)	0859	3.576	0.0907	3.560	0.1120	0.0390	25	0.14	129.00	0
129	Leucine (%)	0098	3.610	0.0030	3.560	0.1120	0.0390	25	0.44	129.02	0
129	Leucine (%)	0353	3.610	0.1000	3.560	0.1120	0.0390	25	0.44	129.00	0
129	Leucine (%)	0571	3.617	0.0230	3.560	0.1120	0.0390	25	0.50	129.00	0
129	Leucine (%)	0226	3.620	0.1400	3.560	0.1120	0.0390	25	0.53	129.00	0
129	Leucine (%)	0504	3.620	0.0000	3.560	0.1120	0.0390	25	0.53	129.00	0
129	Leucine (%)	0682	3.629	0.0000	3.560	0.1120	0.0390	25	0.61	129.00	0
129	Leucine (%)	2188	3.639	0.0000	3.560	0.1120	0.0390	25	0.70	129.05	0
129	Leucine (%)	0354	3.662	0.0050	3.560	0.1120	0.0390	25	0.90	129.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
129	Leucine (%)	0644	3.710	0.0050	3.560	0.1120	0.0390	25	1.33	129.00	0
129	Leucine (%)	0675	3.730	0.0200	3.560	0.1120	0.0390	25	1.51	129.00	0
129	Leucine (%)	0227	3.780	0.0200	3.560	0.1120	0.0390	25	1.96	129.99	0
129	Leucine (%)	0148	3.525	0.3700	3.560	0.1120	0.0390	25	-0.32	129.05	1
130	L-Lysine (%)	0407	2.352	0.0130	2.900	0.1505	0.0431	25	-3.65	130.05	0
130	L-Lysine (%)	0682	2.637	0.0000	2.900	0.1505	0.0431	25	-1.75	130.00	0
130	L-Lysine (%)	2188	2.731	0.0000	2.900	0.1505	0.0431	25	-1.12	130.05	0
130	L-Lysine (%)	2246	2.735	0.0300	2.900	0.1505	0.0431	25	-1.10	130.05	0
130	L-Lysine (%)	0353	2.755	0.0100	2.900	0.1505	0.0431	25	-0.96	130.00	0
130	L-Lysine (%)	0652	2.780	0.0200	2.900	0.1505	0.0431	25	-0.80	130.00	0
130	L-Lysine (%)	0226	2.795	0.1100	2.900	0.1505	0.0431	25	-0.70	130.00	0
130	L-Lysine (%)	0910	2.850	0.0200	2.900	0.1505	0.0431	25	-0.33	130.00	0
130	L-Lysine (%)	0148	2.855	0.2300	2.900	0.1505	0.0431	25	-0.30	130.05	0
130	L-Lysine (%)	0354	2.862	0.0510	2.900	0.1505	0.0431	25	-0.26	130.00	0
130	L-Lysine (%)	0684	2.870	0.0200	2.900	0.1505	0.0431	25	-0.20	130.00	0
130	L-Lysine (%)	0941	2.900	0.0000	2.900	0.1505	0.0431	25	0.00	130.00	0
130	L-Lysine (%)	0968	2.900	0.0000	2.900	0.1505	0.0431	25	0.00	130.00	0
130	L-Lysine (%)	0675	2.920	0.0600	2.900	0.1505	0.0431	25	0.13	130.00	0
130	L-Lysine (%)	0723	2.920	0.0200	2.900	0.1505	0.0431	25	0.13	130.00	0
130	L-Lysine (%)	2059	2.921	0.0020	2.900	0.1505	0.0431	25	0.14	130.00	0
130	L-Lysine (%)	0859	2.947	0.0515	2.900	0.1505	0.0431	25	0.31	130.00	0
130	L-Lysine (%)	0644	2.972	0.0100	2.900	0.1505	0.0431	25	0.48	130.00	0
130	L-Lysine (%)	0571	2.981	0.0380	2.900	0.1505	0.0431	25	0.54	130.00	0
130	L-Lysine (%)	0504	3.020	0.0200	2.900	0.1505	0.0431	25	0.80	130.00	0
130	L-Lysine (%)	0098	3.032	0.0330	2.900	0.1505	0.0431	25	0.87	130.02	0
130	L-Lysine (%)	0918	3.089	0.0700	2.900	0.1505	0.0431	25	1.25	130.00	0
130	L-Lysine (%)	0227	3.095	0.1500	2.900	0.1505	0.0431	25	1.29	130.99	0
130	L-Lysine (%)	2319	3.100	0.0400	2.900	0.1505	0.0431	25	1.33	130.05	0
130	L-Lysine (%)	0676	3.530	0.0800	2.900	0.1505	0.0431	25	4.19	130.05	0
130	L-Lysine (%)	0242	2.675	0.3500	2.900	0.1505	0.0431	25	-1.50	130.05	1
131	Methionine (%)	2246	0.4900	0.0000	0.6416	0.0508	0.0123	25	-2.98	131.05	0
131	Methionine (%)	0941	0.5650	0.0040	0.6416	0.0508	0.0123	25	-1.51	131.00	0
131	Methionine (%)	0354	0.5840	0.0040	0.6416	0.0508	0.0123	25	-1.13	131.00	0
131	Methionine (%)	0918	0.5900	0.0400	0.6416	0.0508	0.0123	25	-1.01	131.00	0
131	Methionine (%)	0148	0.5925	0.0650	0.6416	0.0508	0.0123	25	-0.97	131.05	0
131	Methionine (%)	0652	0.6050	0.0100	0.6416	0.0508	0.0123	25	-0.72	131.00	0
131	Methionine (%)	0353	0.6150	0.0100	0.6416	0.0508	0.0123	25	-0.52	131.00	0
131	Methionine (%)	0684	0.6150	0.0100	0.6416	0.0508	0.0123	25	-0.52	131.00	0
131	Methionine (%)	0723	0.6200	0.0200	0.6416	0.0508	0.0123	25	-0.42	131.00	0
131	Methionine (%)	2188	0.6220	0.0000	0.6416	0.0508	0.0123	25	-0.39	131.05	0
131	Methionine (%)	0226	0.6300	0.0000	0.6416	0.0508	0.0123	25	-0.23	131.00	0
131	Methionine (%)	0242	0.6350	0.0100	0.6416	0.0508	0.0123	25	-0.13	131.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
131	Methionine (%)	0859	0.6415	0.0125	0.6416	0.0508	0.0123	25	0.00	131.00	0
131	Methionine (%)	2059	0.6540	0.0080	0.6416	0.0508	0.0123	25	0.24	131.00	0
131	Methionine (%)	0571	0.6550	0.0040	0.6416	0.0508	0.0123	25	0.26	131.00	0
131	Methionine (%)	0098	0.6565	0.0090	0.6416	0.0508	0.0123	25	0.29	131.02	0
131	Methionine (%)	0504	0.6600	0.0000	0.6416	0.0508	0.0123	25	0.36	131.00	0
131	Methionine (%)	2319	0.6600	0.0000	0.6416	0.0508	0.0123	25	0.36	131.05	0
131	Methionine (%)	0682	0.6730	0.0000	0.6416	0.0508	0.0123	25	0.62	131.00	0
131	Methionine (%)	0644	0.6750	0.0000	0.6416	0.0508	0.0123	25	0.66	131.00	0
131	Methionine (%)	0910	0.6800	0.0200	0.6416	0.0508	0.0123	25	0.76	131.00	0
131	Methionine (%)	0968	0.7000	0.0000	0.6416	0.0508	0.0123	25	1.15	131.00	0
131	Methionine (%)	0227	0.7100	0.0000	0.6416	0.0508	0.0123	25	1.35	131.99	0
131	Methionine (%)	0676	0.7850	0.0100	0.6416	0.0508	0.0123	25	2.82	131.05	0
131	Methionine (%)	0675	0.8150	0.0700	0.6416	0.0508	0.0123	25	3.41	131.00	0
131	Methionine (%)	0407	0.5241	0.1590	0.6416	0.0508	0.0123	25	-2.31	131.05	1
132	Phenylalanine (%)	0682	1.992	0.0000	2.377	0.0983	0.0451	26	-3.92	132.00	0
132	Phenylalanine (%)	0226	2.260	0.1000	2.377	0.0983	0.0451	26	-1.19	132.00	0
132	Phenylalanine (%)	2319	2.270	0.0000	2.377	0.0983	0.0451	26	-1.09	132.05	0
132	Phenylalanine (%)	0948	2.279	0.2510	2.377	0.0983	0.0451	26	-1.00	132.01	0
132	Phenylalanine (%)	0968	2.290	0.0400	2.377	0.0983	0.0451	26	-0.89	132.00	0
132	Phenylalanine (%)	0676	2.290	0.0400	2.377	0.0983	0.0451	26	-0.89	132.05	0
132	Phenylalanine (%)	0684	2.310	0.0800	2.377	0.0983	0.0451	26	-0.68	132.00	0
132	Phenylalanine (%)	0723	2.315	0.0100	2.377	0.0983	0.0451	26	-0.63	132.00	0
132	Phenylalanine (%)	0675	2.325	0.0300	2.377	0.0983	0.0451	26	-0.53	132.00	0
132	Phenylalanine (%)	0652	2.330	0.0200	2.377	0.0983	0.0451	26	-0.48	132.00	0
132	Phenylalanine (%)	0910	2.335	0.0300	2.377	0.0983	0.0451	26	-0.43	132.00	0
132	Phenylalanine (%)	0098	2.337	0.0070	2.377	0.0983	0.0451	26	-0.41	132.02	0
132	Phenylalanine (%)	0354	2.356	0.0220	2.377	0.0983	0.0451	26	-0.22	132.00	0
132	Phenylalanine (%)	2059	2.369	0.0030	2.377	0.0983	0.0451	26	-0.09	132.00	0
132	Phenylalanine (%)	0941	2.399	0.1450	2.377	0.0983	0.0451	26	0.22	132.00	0
132	Phenylalanine (%)	2246	2.405	0.1300	2.377	0.0983	0.0451	26	0.28	132.05	0
132	Phenylalanine (%)	0504	2.425	0.0100	2.377	0.0983	0.0451	26	0.49	132.00	0
132	Phenylalanine (%)	0242	2.440	0.0600	2.377	0.0983	0.0451	26	0.64	132.05	0
132	Phenylalanine (%)	0571	2.446	0.0440	2.377	0.0983	0.0451	26	0.70	132.00	0
132	Phenylalanine (%)	0918	2.451	0.0060	2.377	0.0983	0.0451	26	0.75	132.00	0
132	Phenylalanine (%)	0644	2.464	0.0010	2.377	0.0983	0.0451	26	0.88	132.00	0
132	Phenylalanine (%)	0227	2.465	0.0100	2.377	0.0983	0.0451	26	0.89	132.99	0
132	Phenylalanine (%)	0353	2.490	0.0400	2.377	0.0983	0.0451	26	1.15	132.00	0
132	Phenylalanine (%)	0859	2.495	0.0498	2.377	0.0983	0.0451	26	1.19	132.00	0
132	Phenylalanine (%)	2188	2.510	0.0000	2.377	0.0983	0.0451	26	1.35	132.05	0
132	Phenylalanine (%)	0407	2.771	0.0430	2.377	0.0983	0.0451	26	4.00	132.05	0
132	Phenylalanine (%)	0148	2.800	0.7000	2.377	0.0983	0.0451	26	4.30	132.05	1
133	Proline (%)	0859	1.131	0.0152	2.382	0.1998	0.0456	24	-6.26	133.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
133	Proline (%)	0682	1.295	0.0000	2.382	0.1998	0.0456	24	-5.44	133.00	0
133	Proline (%)	2246	2.160	0.0000	2.382	0.1998	0.0456	24	-1.11	133.05	0
133	Proline (%)	0652	2.200	0.0400	2.382	0.1998	0.0456	24	-0.91	133.00	0
133	Proline (%)	0407	2.209	0.0960	2.382	0.1998	0.0456	24	-0.87	133.05	0
133	Proline (%)	0504	2.225	0.0100	2.382	0.1998	0.0456	24	-0.79	133.00	0
133	Proline (%)	0941	2.284	0.0600	2.382	0.1998	0.0456	24	-0.49	133.00	0
133	Proline (%)	2059	2.304	0.0090	2.382	0.1998	0.0456	24	-0.39	133.00	0
133	Proline (%)	0242	2.305	0.0900	2.382	0.1998	0.0456	24	-0.39	133.05	0
133	Proline (%)	0148	2.320	0.2000	2.382	0.1998	0.0456	24	-0.31	133.05	0
133	Proline (%)	0684	2.355	0.0100	2.382	0.1998	0.0456	24	-0.14	133.00	0
133	Proline (%)	0227	2.365	0.1100	2.382	0.1998	0.0456	24	-0.09	133.99	0
133	Proline (%)	0723	2.380	0.0400	2.382	0.1998	0.0456	24	-0.01	133.00	0
133	Proline (%)	0226	2.410	0.0600	2.382	0.1998	0.0456	24	0.14	133.00	0
133	Proline (%)	2188	2.435	0.0000	2.382	0.1998	0.0456	24	0.26	133.05	0
133	Proline (%)	0644	2.442	0.0060	2.382	0.1998	0.0456	24	0.30	133.00	0
133	Proline (%)	0354	2.446	0.0670	2.382	0.1998	0.0456	24	0.32	133.00	0
133	Proline (%)	0571	2.471	0.0090	2.382	0.1998	0.0456	24	0.44	133.00	0
133	Proline (%)	2319	2.505	0.0100	2.382	0.1998	0.0456	24	0.61	133.05	0
133	Proline (%)	0353	2.600	0.0200	2.382	0.1998	0.0456	24	1.09	133.00	0
133	Proline (%)	0676	2.610	0.0800	2.382	0.1998	0.0456	24	1.14	133.05	0
133	Proline (%)	0675	2.630	0.0200	2.382	0.1998	0.0456	24	1.24	133.00	0
133	Proline (%)	0968	2.675	0.0500	2.382	0.1998	0.0456	24	1.46	133.00	0
133	Proline (%)	0918	3.966	0.0930	2.382	0.1998	0.0456	24	7.93	133.00	0
133	Proline (%)	0910	1.685	1.210	2.382	0.1998	0.0456	24	-3.49	133.00	1
134	Serine (%)	0407	1.825	0.0360	2.344	0.1700	0.0288	26	-3.05	134.05	0
134	Serine (%)	0504	1.960	0.0200	2.344	0.1700	0.0288	26	-2.26	134.00	0
134	Serine (%)	0675	2.000	0.0200	2.344	0.1700	0.0288	26	-2.02	134.00	0
134	Serine (%)	0676	2.030	0.0200	2.344	0.1700	0.0288	26	-1.85	134.05	0
134	Serine (%)	0918	2.208	0.0460	2.344	0.1700	0.0288	26	-0.80	134.00	0
134	Serine (%)	0098	2.234	0.0020	2.344	0.1700	0.0288	26	-0.65	134.02	0
134	Serine (%)	0644	2.248	0.0140	2.344	0.1700	0.0288	26	-0.57	134.00	0
134	Serine (%)	0941	2.262	0.0160	2.344	0.1700	0.0288	26	-0.48	134.00	0
134	Serine (%)	0684	2.295	0.0500	2.344	0.1700	0.0288	26	-0.29	134.00	0
134	Serine (%)	2188	2.313	0.0000	2.344	0.1700	0.0288	26	-0.18	134.05	0
134	Serine (%)	0652	2.320	0.0200	2.344	0.1700	0.0288	26	-0.14	134.00	0
134	Serine (%)	0910	2.350	0.0000	2.344	0.1700	0.0288	26	0.03	134.00	0
134	Serine (%)	0242	2.350	0.0400	2.344	0.1700	0.0288	26	0.03	134.05	0
134	Serine (%)	0148	2.370	0.1200	2.344	0.1700	0.0288	26	0.15	134.05	0
134	Serine (%)	0968	2.375	0.0100	2.344	0.1700	0.0288	26	0.18	134.00	0
134	Serine (%)	2059	2.377	0.0080	2.344	0.1700	0.0288	26	0.19	134.00	0
134	Serine (%)	0571	2.384	0.0230	2.344	0.1700	0.0288	26	0.23	134.00	0
134	Serine (%)	0354	2.415	0.0270	2.344	0.1700	0.0288	26	0.41	134.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
134	Serine (%)	0859	2.431	0.0459	2.344	0.1700	0.0288	26	0.51	134.00	0
134	Serine (%)	0682	2.480	0.0000	2.344	0.1700	0.0288	26	0.80	134.00	0
134	Serine (%)	0227	2.480	0.0200	2.344	0.1700	0.0288	26	0.80	134.99	0
134	Serine (%)	0353	2.495	0.0500	2.344	0.1700	0.0288	26	0.89	134.00	0
134	Serine (%)	0723	2.525	0.0100	2.344	0.1700	0.0288	26	1.06	134.00	0
134	Serine (%)	2246	2.530	0.0600	2.344	0.1700	0.0288	26	1.09	134.05	0
134	Serine (%)	0226	2.550	0.0800	2.344	0.1700	0.0288	26	1.21	134.00	0
134	Serine (%)	2319	2.645	0.0100	2.344	0.1700	0.0288	26	1.77	134.05	0
134	Serine (%)	0948	2.291	0.3080	2.344	0.1700	0.0288	26	-0.31	134.01	1
135	Threonine (%)	0407	1.497	0.0490	1.858	0.0893	0.0267	26	-4.05	135.05	0
135	Threonine (%)	0675	1.715	0.0100	1.858	0.0893	0.0267	26	-1.60	135.00	0
135	Threonine (%)	0676	1.730	0.0000	1.858	0.0893	0.0267	26	-1.43	135.05	0
135	Threonine (%)	0226	1.755	0.0700	1.858	0.0893	0.0267	26	-1.15	135.00	0
135	Threonine (%)	2188	1.787	0.0000	1.858	0.0893	0.0267	26	-0.79	135.05	0
135	Threonine (%)	0504	1.795	0.0100	1.858	0.0893	0.0267	26	-0.70	135.00	0
135	Threonine (%)	0652	1.795	0.0100	1.858	0.0893	0.0267	26	-0.70	135.00	0
135	Threonine (%)	0910	1.810	0.0600	1.858	0.0893	0.0267	26	-0.54	135.00	0
135	Threonine (%)	0644	1.833	0.0020	1.858	0.0893	0.0267	26	-0.28	135.00	0
135	Threonine (%)	0098	1.834	0.0310	1.858	0.0893	0.0267	26	-0.27	135.02	0
135	Threonine (%)	0148	1.845	0.1500	1.858	0.0893	0.0267	26	-0.14	135.05	0
135	Threonine (%)	0684	1.850	0.0000	1.858	0.0893	0.0267	26	-0.09	135.00	0
135	Threonine (%)	0242	1.850	0.1000	1.858	0.0893	0.0267	26	-0.09	135.05	0
135	Threonine (%)	0941	1.852	0.0000	1.858	0.0893	0.0267	26	-0.06	135.00	0
135	Threonine (%)	2059	1.855	0.0020	1.858	0.0893	0.0267	26	-0.03	135.00	0
135	Threonine (%)	0571	1.880	0.0180	1.858	0.0893	0.0267	26	0.25	135.00	0
135	Threonine (%)	0354	1.904	0.0050	1.858	0.0893	0.0267	26	0.51	135.00	0
135	Threonine (%)	0859	1.909	0.0413	1.858	0.0893	0.0267	26	0.57	135.00	0
135	Threonine (%)	0682	1.914	0.0000	1.858	0.0893	0.0267	26	0.63	135.00	0
135	Threonine (%)	2319	1.915	0.0100	1.858	0.0893	0.0267	26	0.64	135.05	0
135	Threonine (%)	0353	1.925	0.0100	1.858	0.0893	0.0267	26	0.75	135.00	0
135	Threonine (%)	0227	1.930	0.0000	1.858	0.0893	0.0267	26	0.81	135.99	0
135	Threonine (%)	0723	1.940	0.0400	1.858	0.0893	0.0267	26	0.92	135.00	0
135	Threonine (%)	0968	1.965	0.0100	1.858	0.0893	0.0267	26	1.20	135.00	0
135	Threonine (%)	2246	2.005	0.0500	1.858	0.0893	0.0267	26	1.65	135.05	0
135	Threonine (%)	0918	2.020	0.0160	1.858	0.0893	0.0267	26	1.82	135.00	0
135	Threonine (%)	0948	1.677	0.3250	1.858	0.0893	0.0267	26	-2.03	135.01	1
136	Tryptophan (%)	2246	0.3600	0.0200	0.6631	0.0803	0.0151	13	-3.77	136.05	0
136	Tryptophan (%)	0682	0.3820	0.0000	0.6631	0.0803	0.0151	13	-3.50	136.00	0
136	Tryptophan (%)	0098	0.6120	0.0180	0.6631	0.0803	0.0151	13	-0.64	136.02	0
136	Tryptophan (%)	0504	0.6450	0.0300	0.6631	0.0803	0.0151	13	-0.23	136.99	0
136	Tryptophan (%)	0859	0.6486	0.0115	0.6631	0.0803	0.0151	13	-0.18	136.03	0
136	Tryptophan (%)	2059	0.6495	0.0010	0.6631	0.0803	0.0151	13	-0.17	136.03	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
136	Tryptophan (%)	0644	0.6600	0.0020	0.6631	0.0803	0.0151	13	-0.04	136.01	0
136	Tryptophan (%)	0571	0.6615	0.0070	0.6631	0.0803	0.0151	13	-0.02	136.03	0
136	Tryptophan (%)	0227	0.7000	0.0000	0.6631	0.0803	0.0151	13	0.46	136.00	0
136	Tryptophan (%)	0918	0.7100	0.0000	0.6631	0.0803	0.0151	13	0.58	136.00	0
136	Tryptophan (%)	0684	0.7250	0.0500	0.6631	0.0803	0.0151	13	0.77	136.00	0
136	Tryptophan (%)	0353	0.7400	0.0400	0.6631	0.0803	0.0151	13	0.96	136.03	0
136	Tryptophan (%)	0407	0.8075	0.0170	0.6631	0.0803	0.0151	13	1.80	136.05	0
137	Tyrosine (%)	0910	1.310	0.0000	1.657	0.1580	0.0524	24	-2.19	137.00	0
137	Tyrosine (%)	2246	1.380	0.0400	1.657	0.1580	0.0524	24	-1.75	137.05	0
137	Tyrosine (%)	0675	1.405	0.0300	1.657	0.1580	0.0524	24	-1.59	137.00	0
137	Tyrosine (%)	0354	1.502	0.0080	1.657	0.1580	0.0524	24	-0.98	137.00	0
137	Tyrosine (%)	0948	1.510	0.2250	1.657	0.1580	0.0524	24	-0.93	137.01	0
137	Tyrosine (%)	2188	1.566	0.0000	1.657	0.1580	0.0524	24	-0.57	137.05	0
137	Tyrosine (%)	0227	1.610	0.0800	1.657	0.1580	0.0524	24	-0.29	137.99	0
137	Tyrosine (%)	0098	1.611	0.0230	1.657	0.1580	0.0524	24	-0.29	137.02	0
137	Tyrosine (%)	0684	1.615	0.0100	1.657	0.1580	0.0524	24	-0.26	137.00	0
137	Tyrosine (%)	0226	1.630	0.0400	1.657	0.1580	0.0524	24	-0.17	137.00	0
137	Tyrosine (%)	0504	1.640	0.0800	1.657	0.1580	0.0524	24	-0.10	137.00	0
137	Tyrosine (%)	0242	1.650	0.1200	1.657	0.1580	0.0524	24	-0.04	137.05	0
137	Tyrosine (%)	0941	1.660	0.2410	1.657	0.1580	0.0524	24	0.02	137.00	0
137	Tyrosine (%)	0644	1.664	0.0060	1.657	0.1580	0.0524	24	0.05	137.00	0
137	Tyrosine (%)	2319	1.670	0.0200	1.657	0.1580	0.0524	24	0.09	137.05	0
137	Tyrosine (%)	0968	1.700	0.0000	1.657	0.1580	0.0524	24	0.28	137.00	0
137	Tyrosine (%)	0682	1.702	0.0000	1.657	0.1580	0.0524	24	0.29	137.00	0
137	Tyrosine (%)	0148	1.765	0.1300	1.657	0.1580	0.0524	24	0.69	137.05	0
137	Tyrosine (%)	0723	1.770	0.0600	1.657	0.1580	0.0524	24	0.72	137.00	0
137	Tyrosine (%)	2059	1.806	0.0040	1.657	0.1580	0.0524	24	0.95	137.00	0
137	Tyrosine (%)	0407	1.812	0.0370	1.657	0.1580	0.0524	24	0.98	137.05	0
137	Tyrosine (%)	0918	1.834	0.0630	1.657	0.1580	0.0524	24	1.12	137.00	0
137	Tyrosine (%)	0676	1.890	0.0400	1.657	0.1580	0.0524	24	1.48	137.05	0
137	Tyrosine (%)	0353	1.910	0.0000	1.657	0.1580	0.0524	24	1.60	137.00	0
138	Valine (%)	0682	1.828	0.0000	2.168	0.1600	0.0372	25	-2.13	138.00	0
138	Valine (%)	2319	1.915	0.0100	2.168	0.1600	0.0372	25	-1.58	138.05	0
138	Valine (%)	0941	1.930	0.0000	2.168	0.1600	0.0372	25	-1.49	138.00	0
138	Valine (%)	0242	2.025	0.2300	2.168	0.1600	0.0372	25	-0.90	138.05	0
138	Valine (%)	0353	2.035	0.0300	2.168	0.1600	0.0372	25	-0.83	138.00	0
138	Valine (%)	2246	2.060	0.0000	2.168	0.1600	0.0372	25	-0.68	138.05	0
138	Valine (%)	0723	2.065	0.0100	2.168	0.1600	0.0372	25	-0.65	138.00	0
138	Valine (%)	0918	2.092	0.0200	2.168	0.1600	0.0372	25	-0.48	138.00	0
138	Valine (%)	2188	2.120	0.0000	2.168	0.1600	0.0372	25	-0.30	138.05	0
138	Valine (%)	0226	2.130	0.1000	2.168	0.1600	0.0372	25	-0.24	138.00	0
138	Valine (%)	0910	2.135	0.1100	2.168	0.1600	0.0372	25	-0.21	138.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
138	Valine (%)	0652	2.145	0.0100	2.168	0.1600	0.0372	25	-0.15	138.00	0
138	Valine (%)	0407	2.169	0.0660	2.168	0.1600	0.0372	25	0.00	138.05	0
138	Valine (%)	0684	2.170	0.0800	2.168	0.1600	0.0372	25	0.01	138.00	0
138	Valine (%)	0354	2.186	0.0080	2.168	0.1600	0.0372	25	0.11	138.00	0
138	Valine (%)	0859	2.218	0.0605	2.168	0.1600	0.0372	25	0.31	138.00	0
138	Valine (%)	2059	2.230	0.0010	2.168	0.1600	0.0372	25	0.38	138.00	0
138	Valine (%)	0098	2.276	0.0490	2.168	0.1600	0.0372	25	0.67	138.02	0
138	Valine (%)	0675	2.285	0.0100	2.168	0.1600	0.0372	25	0.73	138.00	0
138	Valine (%)	0968	2.300	0.0200	2.168	0.1600	0.0372	25	0.82	138.00	0
138	Valine (%)	0676	2.305	0.0300	2.168	0.1600	0.0372	25	0.85	138.05	0
138	Valine (%)	0571	2.324	0.0690	2.168	0.1600	0.0372	25	0.97	138.00	0
138	Valine (%)	0504	2.350	0.0000	2.168	0.1600	0.0372	25	1.14	138.00	0
138	Valine (%)	0227	2.395	0.0100	2.168	0.1600	0.0372	25	1.42	138.99	0
138	Valine (%)	0644	2.434	0.0070	2.168	0.1600	0.0372	25	1.66	138.00	0
138	Valine (%)	0148	2.295	0.3100	2.168	0.1600	0.0372	25	0.79	138.05	1
139	Taurine (%)	0407	0.0165	0.0010				3		139.05	0
139	Taurine (%)	0682	0.0360	0.0000				3		139.00	0
139	Taurine (%)	0504	0.1300	0.0000				3		139.00	0
139	Taurine (%)	0098	< 0.01					3		139.02	5
139	Taurine (%)	0227	< 0.01					3		139.99	5
160	Fructose (%)	0297	0.0400	0.0020				2		160.10	0
160	Fructose (%)	0227	0.4950	0.0100				2		160.99	0
161	Galactose (%)	0297	0.0000	0.0000				0		161.10	4
162	Glucose (%)	0297	0.0070	0.0000				1		162.10	0
162	Glucose (%)	0227	< 0.15					1		162.99	5
163	Lactose (%)	0297	0.0000	0.0000				0		163.10	4
163	Lactose (%)	0227	< 0.15					0		163.99	5
164	Maltose (%)	0297	0.0170	0.0020				2		164.10	0
164	Maltose (%)	0227	0.3350	0.0300				2		164.99	0
165	Sucrose (%)	0297	5.758	0.6640				2		165.10	0
165	Sucrose (%)	0227	7.520	0.0800				2		165.99	0
166	Raffinose (%)	0227	0.9400	0.0000				2		166.99	0
166	Raffinose (%)	0297	1.069	0.0300				2		166.10	0
167	Stachyose (%)	0297	4.724	0.5130				2		167.10	0
167	Stachyose (%)	0227	5.220	0.0600				2		167.99	0
400	Water Activity (Units)	2193	0.6000	0.0000	0.6243	0.0076	0.0037	12	-3.19	400.01	0
400	Water Activity (Units)	2303	0.6135	0.0010	0.6243	0.0076	0.0037	12	-1.42	400.01	0
400	Water Activity (Units)	2268	0.6175	0.0090	0.6243	0.0076	0.0037	12	-0.89	400.01	0
400	Water Activity (Units)	2181	0.6215	0.0008	0.6243	0.0076	0.0037	12	-0.37	400.01	0
400	Water Activity (Units)	0853	0.6224	0.0012	0.6243	0.0076	0.0037	12	-0.25	400.01	0
400	Water Activity (Units)	2174	0.6265	0.0010	0.6243	0.0076	0.0037	12	0.29	400.01	0
400	Water Activity (Units)	0589	0.6270	0.0020	0.6243	0.0076	0.0037	12	0.36	400.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
400	Water Activity (Units)	2168	0.6275	0.0010	0.6243	0.0076	0.0037	12	0.42	400.01	0
400	Water Activity (Units)	2073	0.6290	0.0140	0.6243	0.0076	0.0037	12	0.62	400.99	0
400	Water Activity (Units)	0123	0.6300	0.0000	0.6243	0.0076	0.0037	12	0.75	400.01	0
400	Water Activity (Units)	0407	0.6317	0.0002	0.6243	0.0076	0.0037	12	0.97	400.01	0
400	Water Activity (Units)	0942	0.6320	0.0140	0.6243	0.0076	0.0037	12	1.01	400.01	0
516	Arsenic, Total (As) (ppm)	0171	0.0100	0.0000	0.0106	0.0010	0.0012	4		516.00	0
516	Arsenic, Total (As) (ppm)	0407	0.0101	0.0018	0.0106	0.0010	0.0012	4		516.53	0
516	Arsenic, Total (As) (ppm)	0553	0.0103	0.0009	0.0106	0.0010	0.0012	4		516.53	0
516	Arsenic, Total (As) (ppm)	0227	0.0120	0.0020	0.0106	0.0010	0.0012	4		516.53	0
516	Arsenic, Total (As) (ppm)	0682	< 10		0.0106	0.0010	0.0012	4		516.43	5
518	Cadmium (ppm)	0407	0.0318	0.0024	0.0383	0.0061	0.0019	5		518.41	0
518	Cadmium (ppm)	0407	0.0351	0.0004	0.0383	0.0061	0.0019	5		518.53	0
518	Cadmium (ppm)	0227	0.0370	0.0020	0.0383	0.0061	0.0019	5		518.53	0
518	Cadmium (ppm)	0553	0.0398	0.0009	0.0383	0.0061	0.0019	5		518.53	0
518	Cadmium (ppm)	0171	0.0480	0.0040	0.0383	0.0061	0.0019	5		518.41	0
518	Cadmium (ppm)	0407	0.0404	0.0128	0.0383	0.0061	0.0019	5		518.43	1
518	Cadmium (ppm)	0186	< 0.05		0.0383	0.0061	0.0019	5		518.52	5
518	Cadmium (ppm)	0682	< 0.2		0.0383	0.0061	0.0019	5		518.43	5
520	Chromium, Total (Cr) (ppm)	0186	0.0990	0.0060	0.4060	0.2962	0.0607	9	-1.04	520.52	0
520	Chromium, Total (Cr) (ppm)	0407	0.1039	0.0294	0.4060	0.2962	0.0607	9	-1.02	520.43	0
520	Chromium, Total (Cr) (ppm)	0553	0.2335	0.1310	0.4060	0.2962	0.0607	9	-0.58	520.53	0
520	Chromium, Total (Cr) (ppm)	0407	0.2452	0.0031	0.4060	0.2962	0.0607	9	-0.54	520.41	0
520	Chromium, Total (Cr) (ppm)	0164	0.3790	0.0420	0.4060	0.2962	0.0607	9	-0.09	520.53	0
520	Chromium, Total (Cr) (ppm)	0407	0.4859	0.0374	0.4060	0.2962	0.0607	9	0.27	520.53	0
520	Chromium, Total (Cr) (ppm)	0693	0.6075	0.0870	0.4060	0.2962	0.0607	9	0.68	520.42	0
520	Chromium, Total (Cr) (ppm)	0682	0.6500	0.0000	0.4060	0.2962	0.0607	9	0.82	520.43	0
520	Chromium, Total (Cr) (ppm)	0045	1.025	0.2100	0.4060	0.2962	0.0607	9	2.09	520.42	0
520	Chromium, Total (Cr) (ppm)	0171	0.2535	0.4150	0.4060	0.2962	0.0607	9	-0.51	520.41	1
526	Lead (ppm)	0227	0.0170	0.0000	0.0301	0.0165	0.0109	6	-0.79	526.53	0
526	Lead (ppm)	0553	0.0171	0.0041	0.0301	0.0165	0.0109	6	-0.79	526.53	0
526	Lead (ppm)	0407	0.0211	0.0024	0.0301	0.0165	0.0109	6	-0.55	526.53	0
526	Lead (ppm)	0171	0.0350	0.0020	0.0301	0.0165	0.0109	6	0.30	526.41	0
526	Lead (ppm)	0407	0.0364	0.0184	0.0301	0.0165	0.0109	6	0.38	526.41	0
526	Lead (ppm)	0407	0.0539	0.0384	0.0301	0.0165	0.0109	6	1.45	526.43	0
526	Lead (ppm)	0682	< 3		0.0301	0.0165	0.0109	6		526.43	5
529	Mercury (ppb)	0407	0.8671	0.5493				1		529.99	0
529	Mercury (ppb)	0227	< 10					1		529.99	5
529	Mercury (ppb)	0682	< 10000					1		529.99	5
539	Nickel (ppm)	0171	3.758	0.0490	4.333	0.3986	0.1183	7	-1.44	539.41	0
539	Nickel (ppm)	0407	3.901	0.0866	4.333	0.3986	0.1183	7	-1.08	539.41	0
539	Nickel (ppm)	0407	4.428	0.0858	4.333	0.3986	0.1183	7	0.24	539.53	0
539	Nickel (ppm)	0186	4.520	0.3800	4.333	0.3986	0.1183	7	0.47	539.52	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
539	Nickel (ppm)	0407	4.521	0.1664	4.333	0.3986	0.1183	7	0.47	539.43	0
539	Nickel (ppm)	0682	4.570	0.0000	4.333	0.3986	0.1183	7	0.60	539.43	0
539	Nickel (ppm)	0553	4.630	0.0600	4.333	0.3986	0.1183	7	0.75	539.53	0
710	Lauric Acid (12:0) (% (w/w))	0226	< 0.005					0		710.99	5
714	Myristic Acid (14:0) (% (w/w))	0226	< 0.005					0		714.99	5
716	Palmitic Acid (16:0) (% (w/w))	0226	0.3000	0.0000				1		716.99	0
718	Palmitoleic Acid (9c-16:1) (% (w/w))	0226	< 0.005					0		718.99	5
722	Stearic Acid (18:0) (% (w/w))	0226	0.0650	0.0100				1		722.99	0
724	Oleic Acid (9c-18:1) (% (w/w))	0226	0.1600	0.0000				1		724.99	0
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0226	0.9850	0.0100				1		726.99	0
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/w))	0226	0.1450	0.0100				1		728.99	0
730	Arachidic Acid (20:0) (% (w/w))	0226	< 0.005					0		730.99	5
732	Gondoic Acid (11c-20:1) (% (w/w))	0226	< 0.005					0		732.99	5
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c) (% (w/w))	0226	< 0.005					0		740.99	5
742	Behenic Acid (22:0) (% (w/w))	0226	0.0100	0.0000				1		742.99	0
744	Erucic Acid (13c-22:1) (% (w/w))	0226	< 0.005					0		744.99	5
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c) (% (w/w))	0226	< 0.005					0		746.99	5
748	Lignoceric Acid (24:0) (% (w/w))	0226	< 0.005					0		748.99	5
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22c) (% (w/w))	0226	< 0.005					0		750.99	5
752	Nervonic Acid (24:1) isomers (% (w/w))	0226	< 0.005					0		752.99	5
772	Total Fatty Acids (% (w/w))	0226	1.660	0.0200				1		772.99	0
912	Starch, Polarimetric (Ewers) Data Extracted (% (w/w))	2193	5.635	0.2300	5.801	0.1522	0.1471	7	-1.09	912.00	0
912	Starch, Polarimetric (Ewers) Data Extracted (% (w/w))	0638	5.700	0.0000	5.801	0.1522	0.1471	7	-0.67	912.00	0
912	Starch, Polarimetric (Ewers) Data Extracted (% (w/w))	0354	5.740	0.0200	5.801	0.1522	0.1471	7	-0.40	912.00	0
912	Starch, Polarimetric (Ewers) Data Extracted (% (w/w))	2188	5.755	0.1300	5.801	0.1522	0.1471	7	-0.30	912.00	0
912	Starch, Polarimetric (Ewers) Data Extracted (% (w/w))	2006	5.850	0.0200	5.801	0.1522	0.1471	7	0.32	912.00	0
912	Starch, Polarimetric (Ewers) Data Extracted (% (w/w))	0164	5.900	0.2000	5.801	0.1522	0.1471	7	0.65	912.00	0
912	Starch, Polarimetric (Ewers) Data Extracted (% (w/w))	2181	8.045	0.4300	5.801	0.1522	0.1471	7	14.74	912.00	0

Notes: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2 and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = Rejected for duplicates too far apart, 2 = Rejected as outlier, 5 = A reporting limit and 4 = zeros submitted as values. Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). Flag 3 indicates not used in statistics.