



AAFCO
Proficiency Testing Program



Animal Feed Scheme

Poultry Feed, Medicated

Test Material Code # 202132

Labs Reporting: 174

Analytes Reported 116

Issue Date : 01/31/2022

Analyte Proficiency Testing Report

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
000	Urea (%)	0278	0.4000	0.0000				1		000.02	0
001	Loss on Drying (%)	0612	7.275	0.3700	9.173	0.2530	0.1315	74	-7.50	001.99	0
001	Loss on Drying (%)	0852	8.165	0.7300	9.173	0.2530	0.1315	74	-3.98	001.99	0
001	Loss on Drying (%)	0536	8.300	0.4000	9.173	0.2530	0.1315	74	-3.45	001.99	0
001	Loss on Drying (%)	0345	8.405	0.0300	9.173	0.2530	0.1315	74	-3.03	001.07	0
001	Loss on Drying (%)	0019	8.450	0.1000	9.173	0.2530	0.1315	74	-2.86	001.07	0
001	Loss on Drying (%)	0889	8.640	0.0200	9.173	0.2530	0.1315	74	-2.11	001.99	0
001	Loss on Drying (%)	0010	8.700	0.0000	9.173	0.2530	0.1315	74	-1.87	001.07	0
001	Loss on Drying (%)	2246	8.750	0.1000	9.173	0.2530	0.1315	74	-1.67	001.99	0
001	Loss on Drying (%)	0034	8.790	0.1000	9.173	0.2530	0.1315	74	-1.51	001.07	0
001	Loss on Drying (%)	0675	8.800	0.1400	9.173	0.2530	0.1315	74	-1.47	001.07	0
001	Loss on Drying (%)	0186	8.900	0.0430	9.173	0.2530	0.1315	74	-1.08	001.99	0
001	Loss on Drying (%)	0049	8.915	0.6500	9.173	0.2530	0.1315	74	-1.02	001.07	0
001	Loss on Drying (%)	0309	8.920	0.0600	9.173	0.2530	0.1315	74	-1.00	001.00	0
001	Loss on Drying (%)	0683	8.925	0.0100	9.173	0.2530	0.1315	74	-0.98	001.07	0
001	Loss on Drying (%)	0353	9.000	0.2200	9.173	0.2530	0.1315	74	-0.68	001.07	0
001	Loss on Drying (%)	0510	9.000	0.0000	9.173	0.2530	0.1315	74	-0.68	001.99	0
001	Loss on Drying (%)	0910	9.000	0.0000	9.173	0.2530	0.1315	74	-0.68	001.99	0
001	Loss on Drying (%)	2144	9.015	0.0100	9.173	0.2530	0.1315	74	-0.62	001.07	0
001	Loss on Drying (%)	0297	9.025	0.5300	9.173	0.2530	0.1315	74	-0.58	001.07	0
001	Loss on Drying (%)	0918	9.060	0.1400	9.173	0.2530	0.1315	74	-0.45	001.99	0
001	Loss on Drying (%)	0650	9.070	0.0200	9.173	0.2530	0.1315	74	-0.41	001.07	0
001	Loss on Drying (%)	0089	9.075	0.0500	9.173	0.2530	0.1315	74	-0.39	001.07	0
001	Loss on Drying (%)	0948	9.080	0.0200	9.173	0.2530	0.1315	74	-0.37	001.99	0
001	Loss on Drying (%)	0083	9.090	0.0600	9.173	0.2530	0.1315	74	-0.33	001.07	0
001	Loss on Drying (%)	0619	9.100	0.0400	9.173	0.2530	0.1315	74	-0.29	001.03	0
001	Loss on Drying (%)	0169	9.105	0.0700	9.173	0.2530	0.1315	74	-0.27	001.00	0
001	Loss on Drying (%)	0643	9.111	0.0010	9.173	0.2530	0.1315	74	-0.25	001.07	0
001	Loss on Drying (%)	2059	9.115	0.0100	9.173	0.2530	0.1315	74	-0.23	001.07	0
001	Loss on Drying (%)	2155	9.125	0.3900	9.173	0.2530	0.1315	74	-0.19	001.99	0
001	Loss on Drying (%)	0682	9.130	0.0000	9.173	0.2530	0.1315	74	-0.17	001.07	0
001	Loss on Drying (%)	0038	9.135	0.1100	9.173	0.2530	0.1315	74	-0.15	001.07	0
001	Loss on Drying (%)	0504	9.135	0.0100	9.173	0.2530	0.1315	74	-0.15	001.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
001	Loss on Drying (%)	2009	9.148	0.0150	9.173	0.2530	0.1315	74	-0.10	001.07	0
001	Loss on Drying (%)	2319	9.150	0.1000	9.173	0.2530	0.1315	74	-0.09	001.99	0
001	Loss on Drying (%)	0644	9.159	0.0270	9.173	0.2530	0.1315	74	-0.06	001.05	0
001	Loss on Drying (%)	0571	9.160	0.0000	9.173	0.2530	0.1315	74	-0.05	001.07	0
001	Loss on Drying (%)	0629	9.190	0.0200	9.173	0.2530	0.1315	74	0.07	001.99	0
001	Loss on Drying (%)	0505	9.195	0.1500	9.173	0.2530	0.1315	74	0.09	001.07	0
001	Loss on Drying (%)	0581	9.200	0.2000	9.173	0.2530	0.1315	74	0.11	001.07	0
001	Loss on Drying (%)	2268	9.215	0.0700	9.173	0.2530	0.1315	74	0.17	001.99	0
001	Loss on Drying (%)	0885	9.215	0.0406	9.173	0.2530	0.1315	74	0.17	001.07	0
001	Loss on Drying (%)	0066	9.216	0.0810	9.173	0.2530	0.1315	74	0.17	001.07	0
001	Loss on Drying (%)	0278	9.220	0.2400	9.173	0.2530	0.1315	74	0.19	001.07	0
001	Loss on Drying (%)	0074	9.230	0.5400	9.173	0.2530	0.1315	74	0.23	001.07	0
001	Loss on Drying (%)	0357	9.230	0.0015	9.173	0.2530	0.1315	74	0.23	001.99	0
001	Loss on Drying (%)	0590	9.245	0.0500	9.173	0.2530	0.1315	74	0.29	001.07	0
001	Loss on Drying (%)	0689	9.250	0.1000	9.173	0.2530	0.1315	74	0.31	001.07	0
001	Loss on Drying (%)	2190	9.250	0.0200	9.173	0.2530	0.1315	74	0.31	001.07	0
001	Loss on Drying (%)	0171	9.260	0.0800	9.173	0.2530	0.1315	74	0.35	001.07	0
001	Loss on Drying (%)	0003	9.270	0.0200	9.173	0.2530	0.1315	74	0.38	001.07	0
001	Loss on Drying (%)	0226	9.280	0.0800	9.173	0.2530	0.1315	74	0.42	001.07	0
001	Loss on Drying (%)	2192	9.285	0.0500	9.173	0.2530	0.1315	74	0.44	001.07	0
001	Loss on Drying (%)	0035	9.300	0.1400	9.173	0.2530	0.1315	74	0.50	001.07	0
001	Loss on Drying (%)	0853	9.300	0.0000	9.173	0.2530	0.1315	74	0.50	001.07	0
001	Loss on Drying (%)	2259	9.318	0.1320	9.173	0.2530	0.1315	74	0.57	001.07	0
001	Loss on Drying (%)	0028	9.320	0.1200	9.173	0.2530	0.1315	74	0.58	001.07	0
001	Loss on Drying (%)	2109	9.320	0.3800	9.173	0.2530	0.1315	74	0.58	001.07	0
001	Loss on Drying (%)	0033	9.335	0.0100	9.173	0.2530	0.1315	74	0.64	001.07	0
001	Loss on Drying (%)	0366	9.350	0.5000	9.173	0.2530	0.1315	74	0.70	001.07	0
001	Loss on Drying (%)	0512	9.355	0.2500	9.173	0.2530	0.1315	74	0.72	001.07	0
001	Loss on Drying (%)	2073	9.365	0.0700	9.173	0.2530	0.1315	74	0.76	001.00	0
001	Loss on Drying (%)	0027	9.381	0.0220	9.173	0.2530	0.1315	74	0.82	001.07	0
001	Loss on Drying (%)	0037	9.415	0.1900	9.173	0.2530	0.1315	74	0.96	001.99	0
001	Loss on Drying (%)	0148	9.450	0.1000	9.173	0.2530	0.1315	74	1.10	001.07	0
001	Loss on Drying (%)	0876	9.450	0.1000	9.173	0.2530	0.1315	74	1.10	001.07	0
001	Loss on Drying (%)	0098	9.470	0.2800	9.173	0.2530	0.1315	74	1.18	001.07	0
001	Loss on Drying (%)	2339	9.505	0.0300	9.173	0.2530	0.1315	74	1.31	001.07	0
001	Loss on Drying (%)	0019	9.520	0.2000	9.173	0.2530	0.1315	74	1.37	001.99	0
001	Loss on Drying (%)	0142	9.530	0.1800	9.173	0.2530	0.1315	74	1.41	001.07	0
001	Loss on Drying (%)	0407	9.566	0.0146	9.173	0.2530	0.1315	74	1.56	001.07	0
001	Loss on Drying (%)	0045	9.605	0.4300	9.173	0.2530	0.1315	74	1.71	001.07	0
001	Loss on Drying (%)	0843	9.615	0.0300	9.173	0.2530	0.1315	74	1.75	001.03	0
001	Loss on Drying (%)	0843	9.615	0.0300	9.173	0.2530	0.1315	74	1.75	001.07	0
001	Loss on Drying (%)	0560	9.700	0.2000	9.173	0.2530	0.1315	74	2.08	001.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
001	Loss on Drying (%)	0693	9.445	1.170	9.173	0.2530	0.1315	74	1.08	001.07	1
002	Protein, Crude (%)	0889	17.24	0.0400	20.12	0.2814	0.1478	165	-10.25	002.11	0
002	Protein, Crude (%)	0852	18.75	0.1000	20.12	0.2814	0.1478	165	-4.88	002.05	0
002	Protein, Crude (%)	0553	19.12	0.3000	20.12	0.2814	0.1478	165	-3.57	002.11	0
002	Protein, Crude (%)	0685	19.14	0.4000	20.12	0.2814	0.1478	165	-3.50	002.01	0
002	Protein, Crude (%)	2109	19.27	0.0400	20.12	0.2814	0.1478	165	-3.04	002.05	0
002	Protein, Crude (%)	0536	19.30	0.4000	20.12	0.2814	0.1478	165	-2.93	002.05	0
002	Protein, Crude (%)	0651	19.53	0.1156	20.12	0.2814	0.1478	165	-2.10	002.05	0
002	Protein, Crude (%)	2246	19.54	0.0600	20.12	0.2814	0.1478	165	-2.08	002.05	0
002	Protein, Crude (%)	0510	19.60	0.0000	20.12	0.2814	0.1478	165	-1.86	002.06	0
002	Protein, Crude (%)	0884	19.60	0.4000	20.12	0.2814	0.1478	165	-1.86	002.06	0
002	Protein, Crude (%)	0852	19.60	0.0200	20.12	0.2814	0.1478	165	-1.86	002.11	0
002	Protein, Crude (%)	2144	19.65	0.3600	20.12	0.2814	0.1478	165	-1.69	002.05	0
002	Protein, Crude (%)	0885	19.65	0.0190	20.12	0.2814	0.1478	165	-1.68	002.05	0
002	Protein, Crude (%)	2246	19.65	0.0744	20.12	0.2814	0.1478	165	-1.67	002.06	0
002	Protein, Crude (%)	2193	19.66	0.0100	20.12	0.2814	0.1478	165	-1.67	002.01	0
002	Protein, Crude (%)	2073	19.69	0.2900	20.12	0.2814	0.1478	165	-1.56	002.05	0
002	Protein, Crude (%)	0263	19.70	0.0060	20.12	0.2814	0.1478	165	-1.50	002.06	0
002	Protein, Crude (%)	2006	19.71	0.2600	20.12	0.2814	0.1478	165	-1.47	002.05	0
002	Protein, Crude (%)	0689	19.75	0.5000	20.12	0.2814	0.1478	165	-1.33	002.05	0
002	Protein, Crude (%)	2009	19.76	0.0021	20.12	0.2814	0.1478	165	-1.30	002.05	0
002	Protein, Crude (%)	0653	19.77	0.1100	20.12	0.2814	0.1478	165	-1.28	002.01	0
002	Protein, Crude (%)	0889	19.77	0.0300	20.12	0.2814	0.1478	165	-1.28	002.01	0
002	Protein, Crude (%)	2268	19.81	0.1900	20.12	0.2814	0.1478	165	-1.13	002.06	0
002	Protein, Crude (%)	0123	19.82	0.1000	20.12	0.2814	0.1478	165	-1.08	002.06	0
002	Protein, Crude (%)	0043	19.82	0.1320	20.12	0.2814	0.1478	165	-1.07	002.01	0
002	Protein, Crude (%)	0968	19.83	0.0100	20.12	0.2814	0.1478	165	-1.06	002.01	0
002	Protein, Crude (%)	0539	19.84	0.0100	20.12	0.2814	0.1478	165	-1.03	002.06	0
002	Protein, Crude (%)	2181	19.84	0.0600	20.12	0.2814	0.1478	165	-1.01	002.05	0
002	Protein, Crude (%)	0598	19.84	0.2200	20.12	0.2814	0.1478	165	-1.01	002.06	0
002	Protein, Crude (%)	0619	19.85	0.1000	20.12	0.2814	0.1478	165	-0.97	002.05	0
002	Protein, Crude (%)	2325	19.86	0.2100	20.12	0.2814	0.1478	165	-0.96	002.01	0
002	Protein, Crude (%)	0870	19.88	0.4560	20.12	0.2814	0.1478	165	-0.87	002.01	0
002	Protein, Crude (%)	0100	19.90	0.2300	20.12	0.2814	0.1478	165	-0.81	002.06	0
002	Protein, Crude (%)	0043	19.90	0.0430	20.12	0.2814	0.1478	165	-0.81	002.06	0
002	Protein, Crude (%)	0910	19.90	0.2000	20.12	0.2814	0.1478	165	-0.80	002.01	0
002	Protein, Crude (%)	2200	19.90	0.2000	20.12	0.2814	0.1478	165	-0.80	002.05	0
002	Protein, Crude (%)	0505	19.91	0.2500	20.12	0.2814	0.1478	165	-0.78	002.06	0
002	Protein, Crude (%)	0881	19.95	0.0580	20.12	0.2814	0.1478	165	-0.64	002.05	0
002	Protein, Crude (%)	2192	19.96	0.0700	20.12	0.2814	0.1478	165	-0.60	002.05	0
002	Protein, Crude (%)	0345	19.96	0.1100	20.12	0.2814	0.1478	165	-0.60	002.06	0
002	Protein, Crude (%)	2184	19.96	0.0300	20.12	0.2814	0.1478	165	-0.60	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	2303	19.96	0.1800	20.12	0.2814	0.1478	165	-0.58	002.06	0
002	Protein, Crude (%)	0164	19.97	0.1800	20.12	0.2814	0.1478	165	-0.55	002.06	0
002	Protein, Crude (%)	0885	19.98	0.0300	20.12	0.2814	0.1478	165	-0.53	002.11	0
002	Protein, Crude (%)	2259	19.98	0.0340	20.12	0.2814	0.1478	165	-0.50	002.01	0
002	Protein, Crude (%)	0354	19.99	0.0100	20.12	0.2814	0.1478	165	-0.50	002.05	0
002	Protein, Crude (%)	2022	20.00	0.0100	20.12	0.2814	0.1478	165	-0.46	002.05	0
002	Protein, Crude (%)	0876	20.00	0.0000	20.12	0.2814	0.1478	165	-0.44	002.06	0
002	Protein, Crude (%)	2188	20.01	0.3100	20.12	0.2814	0.1478	165	-0.42	002.01	0
002	Protein, Crude (%)	0066	20.01	0.1200	20.12	0.2814	0.1478	165	-0.41	002.02	0
002	Protein, Crude (%)	0504	20.01	0.1600	20.12	0.2814	0.1478	165	-0.41	002.04	0
002	Protein, Crude (%)	0504	20.01	0.1800	20.12	0.2814	0.1478	165	-0.41	002.06	0
002	Protein, Crude (%)	2146	20.03	0.1300	20.12	0.2814	0.1478	165	-0.35	002.05	0
002	Protein, Crude (%)	0229	20.03	0.3600	20.12	0.2814	0.1478	165	-0.34	002.06	0
002	Protein, Crude (%)	0643	20.04	0.0700	20.12	0.2814	0.1478	165	-0.32	002.05	0
002	Protein, Crude (%)	2340	20.05	0.0010	20.12	0.2814	0.1478	165	-0.27	002.05	0
002	Protein, Crude (%)	0626	20.05	0.3260	20.12	0.2814	0.1478	165	-0.27	002.06	0
002	Protein, Crude (%)	0726	20.05	0.0600	20.12	0.2814	0.1478	165	-0.26	002.06	0
002	Protein, Crude (%)	0001	20.06	0.2900	20.12	0.2814	0.1478	165	-0.25	002.06	0
002	Protein, Crude (%)	0914	20.06	0.0700	20.12	0.2814	0.1478	165	-0.25	002.06	0
002	Protein, Crude (%)	2196	20.06	0.0000	20.12	0.2814	0.1478	165	-0.23	002.01	0
002	Protein, Crude (%)	0009	20.06	0.1400	20.12	0.2814	0.1478	165	-0.23	002.06	0
002	Protein, Crude (%)	0297	20.06	0.1800	20.12	0.2814	0.1478	165	-0.23	002.06	0
002	Protein, Crude (%)	0693	20.06	0.1250	20.12	0.2814	0.1478	165	-0.22	002.06	0
002	Protein, Crude (%)	0190	20.07	0.0300	20.12	0.2814	0.1478	165	-0.21	002.06	0
002	Protein, Crude (%)	0529	20.07	0.0700	20.12	0.2814	0.1478	165	-0.21	002.06	0
002	Protein, Crude (%)	0047	20.09	0.0500	20.12	0.2814	0.1478	165	-0.14	002.06	0
002	Protein, Crude (%)	0644	20.09	0.0940	20.12	0.2814	0.1478	165	-0.13	002.06	0
002	Protein, Crude (%)	0674	20.09	0.0000	20.12	0.2814	0.1478	165	-0.12	002.01	0
002	Protein, Crude (%)	0144	20.09	0.2000	20.12	0.2814	0.1478	165	-0.12	002.06	0
002	Protein, Crude (%)	0870	20.09	0.0838	20.12	0.2814	0.1478	165	-0.11	002.06	0
002	Protein, Crude (%)	2199	20.10	0.4000	20.12	0.2814	0.1478	165	-0.09	002.05	0
002	Protein, Crude (%)	0278	20.10	0.0000	20.12	0.2814	0.1478	165	-0.09	002.06	0
002	Protein, Crude (%)	2319	20.10	0.0000	20.12	0.2814	0.1478	165	-0.09	002.06	0
002	Protein, Crude (%)	0590	20.11	0.0500	20.12	0.2814	0.1478	165	-0.07	002.06	0
002	Protein, Crude (%)	0027	20.11	0.0940	20.12	0.2814	0.1478	165	-0.06	002.06	0
002	Protein, Crude (%)	0169	20.11	0.3000	20.12	0.2814	0.1478	165	-0.05	002.00	0
002	Protein, Crude (%)	0037	20.11	0.2200	20.12	0.2814	0.1478	165	-0.05	002.06	0
002	Protein, Crude (%)	0353	20.11	0.0200	20.12	0.2814	0.1478	165	-0.05	002.06	0
002	Protein, Crude (%)	0407	20.11	0.0090	20.12	0.2814	0.1478	165	-0.04	002.06	0
002	Protein, Crude (%)	0142	20.13	0.0900	20.12	0.2814	0.1478	165	0.00	002.06	0
002	Protein, Crude (%)	0687	20.13	0.6500	20.12	0.2814	0.1478	165	0.00	002.06	0
002	Protein, Crude (%)	0948	20.13	0.1100	20.12	0.2814	0.1478	165	0.00	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	2168	20.13	0.0300	20.12	0.2814	0.1478	165	0.00	002.06	0
002	Protein, Crude (%)	0309	20.14	0.0920	20.12	0.2814	0.1478	165	0.04	002.06	0
002	Protein, Crude (%)	0546	20.14	0.3710	20.12	0.2814	0.1478	165	0.06	002.06	0
002	Protein, Crude (%)	0033	20.15	0.1000	20.12	0.2814	0.1478	165	0.09	002.06	0
002	Protein, Crude (%)	0042	20.15	0.1000	20.12	0.2814	0.1478	165	0.09	002.06	0
002	Protein, Crude (%)	0366	20.15	0.3000	20.12	0.2814	0.1478	165	0.09	002.06	0
002	Protein, Crude (%)	0910	20.15	0.3000	20.12	0.2814	0.1478	165	0.09	002.06	0
002	Protein, Crude (%)	0948	20.16	0.1700	20.12	0.2814	0.1478	165	0.11	002.05	0
002	Protein, Crude (%)	0520	20.16	0.1900	20.12	0.2814	0.1478	165	0.11	002.06	0
002	Protein, Crude (%)	2129	20.17	0.0210	20.12	0.2814	0.1478	165	0.15	002.06	0
002	Protein, Crude (%)	0563	20.17	0.2563	20.12	0.2814	0.1478	165	0.16	002.08	0
002	Protein, Crude (%)	2290	20.19	0.0500	20.12	0.2814	0.1478	165	0.22	002.11	0
002	Protein, Crude (%)	0859	20.19	0.0190	20.12	0.2814	0.1478	165	0.23	002.06	0
002	Protein, Crude (%)	2337	20.20	0.0100	20.12	0.2814	0.1478	165	0.25	002.05	0
002	Protein, Crude (%)	0035	20.20	0.1900	20.12	0.2814	0.1478	165	0.25	002.06	0
002	Protein, Crude (%)	2359	20.20	0.0000	20.12	0.2814	0.1478	165	0.27	002.06	0
002	Protein, Crude (%)	2190	20.20	0.1200	20.12	0.2814	0.1478	165	0.27	002.05	0
002	Protein, Crude (%)	0354	20.20	0.0200	20.12	0.2814	0.1478	165	0.27	002.06	0
002	Protein, Crude (%)	0511	20.20	0.1600	20.12	0.2814	0.1478	165	0.27	002.06	0
002	Protein, Crude (%)	2174	20.20	0.3000	20.12	0.2814	0.1478	165	0.27	002.06	0
002	Protein, Crude (%)	0553	20.20	0.1670	20.12	0.2814	0.1478	165	0.28	002.06	0
002	Protein, Crude (%)	0683	20.21	0.0500	20.12	0.2814	0.1478	165	0.29	002.05	0
002	Protein, Crude (%)	0004	20.22	0.3500	20.12	0.2814	0.1478	165	0.32	002.06	0
002	Protein, Crude (%)	0171	20.22	0.1700	20.12	0.2814	0.1478	165	0.32	002.06	0
002	Protein, Crude (%)	0013	20.23	0.1900	20.12	0.2814	0.1478	165	0.36	002.06	0
002	Protein, Crude (%)	0939	20.23	0.1600	20.12	0.2814	0.1478	165	0.38	002.05	0
002	Protein, Crude (%)	0541	20.23	0.1400	20.12	0.2814	0.1478	165	0.38	002.06	0
002	Protein, Crude (%)	0968	20.24	0.0200	20.12	0.2814	0.1478	165	0.41	002.06	0
002	Protein, Crude (%)	0098	20.25	0.1300	20.12	0.2814	0.1478	165	0.43	002.08	0
002	Protein, Crude (%)	0010	20.25	0.1000	20.12	0.2814	0.1478	165	0.45	002.06	0
002	Protein, Crude (%)	0016	20.25	0.1000	20.12	0.2814	0.1478	165	0.45	002.06	0
002	Protein, Crude (%)	0019	20.25	0.1000	20.12	0.2814	0.1478	165	0.45	002.06	0
002	Protein, Crude (%)	0619	20.25	0.1000	20.12	0.2814	0.1478	165	0.45	002.06	0
002	Protein, Crude (%)	0148	20.26	0.0100	20.12	0.2814	0.1478	165	0.46	002.06	0
002	Protein, Crude (%)	0108	20.26	0.0200	20.12	0.2814	0.1478	165	0.48	002.06	0
002	Protein, Crude (%)	0646	20.26	0.0200	20.12	0.2814	0.1478	165	0.48	002.06	0
002	Protein, Crude (%)	0964	20.28	0.1250	20.12	0.2814	0.1478	165	0.55	002.06	0
002	Protein, Crude (%)	2059	20.28	0.2281	20.12	0.2814	0.1478	165	0.55	002.06	0
002	Protein, Crude (%)	0164	20.28	0.0600	20.12	0.2814	0.1478	165	0.55	002.01	0
002	Protein, Crude (%)	0298	20.28	0.0200	20.12	0.2814	0.1478	165	0.55	002.06	0
002	Protein, Crude (%)	0512	20.29	0.0600	20.12	0.2814	0.1478	165	0.59	002.06	0
002	Protein, Crude (%)	0098	20.30	0.2000	20.12	0.2814	0.1478	165	0.62	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0652	20.30	0.0000	20.12	0.2814	0.1478	165	0.62	002.01	0
002	Protein, Crude (%)	0265	20.30	0.0000	20.12	0.2814	0.1478	165	0.62	002.06	0
002	Protein, Crude (%)	0650	20.30	0.6000	20.12	0.2814	0.1478	165	0.62	002.06	0
002	Protein, Crude (%)	0956	20.30	0.0000	20.12	0.2814	0.1478	165	0.62	002.06	0
002	Protein, Crude (%)	0357	20.31	0.0000	20.12	0.2814	0.1478	165	0.67	002.06	0
002	Protein, Crude (%)	0425	20.32	0.1300	20.12	0.2814	0.1478	165	0.68	002.06	0
002	Protein, Crude (%)	0574	20.34	0.0500	20.12	0.2814	0.1478	165	0.75	002.06	0
002	Protein, Crude (%)	0011	20.34	0.1600	20.12	0.2814	0.1478	165	0.77	002.06	0
002	Protein, Crude (%)	0045	20.35	0.3000	20.12	0.2814	0.1478	165	0.80	002.06	0
002	Protein, Crude (%)	0028	20.38	0.1100	20.12	0.2814	0.1478	165	0.89	002.06	0
002	Protein, Crude (%)	0675	20.38	0.1300	20.12	0.2814	0.1478	165	0.89	002.06	0
002	Protein, Crude (%)	0036	20.39	0.1290	20.12	0.2814	0.1478	165	0.94	002.06	0
002	Protein, Crude (%)	0062	20.39	0.1000	20.12	0.2814	0.1478	165	0.94	002.06	0
002	Protein, Crude (%)	0017	20.40	0.2000	20.12	0.2814	0.1478	165	0.98	002.06	0
002	Protein, Crude (%)	0629	20.41	0.0500	20.12	0.2814	0.1478	165	1.00	002.10	0
002	Protein, Crude (%)	2339	20.42	0.0600	20.12	0.2814	0.1478	165	1.05	002.06	0
002	Protein, Crude (%)	0034	20.43	0.1320	20.12	0.2814	0.1478	165	1.08	002.06	0
002	Protein, Crude (%)	0843	20.44	0.4000	20.12	0.2814	0.1478	165	1.12	002.06	0
002	Protein, Crude (%)	0848	20.44	0.0600	20.12	0.2814	0.1478	165	1.12	002.06	0
002	Protein, Crude (%)	0074	20.45	0.0500	20.12	0.2814	0.1478	165	1.14	002.06	0
002	Protein, Crude (%)	0049	20.45	0.3000	20.12	0.2814	0.1478	165	1.16	002.06	0
002	Protein, Crude (%)	0003	20.48	0.8100	20.12	0.2814	0.1478	165	1.25	002.06	0
002	Protein, Crude (%)	0358	20.48	0.3100	20.12	0.2814	0.1478	165	1.25	002.06	0
002	Protein, Crude (%)	0226	20.49	0.0500	20.12	0.2814	0.1478	165	1.28	002.06	0
002	Protein, Crude (%)	0571	20.49	0.0800	20.12	0.2814	0.1478	165	1.30	002.06	0
002	Protein, Crude (%)	0018	20.51	0.5600	20.12	0.2814	0.1478	165	1.37	002.06	0
002	Protein, Crude (%)	0083	20.53	0.0700	20.12	0.2814	0.1478	165	1.42	002.06	0
002	Protein, Crude (%)	0692	20.53	0.1400	20.12	0.2814	0.1478	165	1.44	002.06	0
002	Protein, Crude (%)	2076	20.54	0.1810	20.12	0.2814	0.1478	165	1.49	002.06	0
002	Protein, Crude (%)	0880	20.55	0.1000	20.12	0.2814	0.1478	165	1.51	002.06	0
002	Protein, Crude (%)	2302	20.58	0.0300	20.12	0.2814	0.1478	165	1.60	002.06	0
002	Protein, Crude (%)	2109	20.59	0.1300	20.12	0.2814	0.1478	165	1.64	002.06	0
002	Protein, Crude (%)	2150	20.62	0.0200	20.12	0.2814	0.1478	165	1.76	002.06	0
002	Protein, Crude (%)	0186	20.68	0.1100	20.12	0.2814	0.1478	165	1.96	002.06	0
002	Protein, Crude (%)	0921	20.74	0.3070	20.12	0.2814	0.1478	165	2.19	002.06	0
002	Protein, Crude (%)	2213	20.81	0.2000	20.12	0.2814	0.1478	165	2.44	002.06	0
002	Protein, Crude (%)	2089	20.82	0.0800	20.12	0.2814	0.1478	165	2.47	002.06	0
002	Protein, Crude (%)	0953	21.08	0.4800	20.12	0.2814	0.1478	165	3.40	002.06	0
002	Protein, Crude (%)	0853	21.20	0.6000	20.12	0.2814	0.1478	165	3.82	002.06	0
002	Protein, Crude (%)	2155	21.27	0.1940	20.12	0.2814	0.1478	165	4.07	002.06	0
002	Protein, Crude (%)	0300	20.10	1.000	20.12	0.2814	0.1478	165	-0.09	002.06	1
002	Protein, Crude (%)	0918	8.910	0.1600	20.12	0.2814	0.1478	165	-39.85	002.06	2

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	2105	33.40	0.2000	20.12	0.2814	0.1478	165	47.17	002.11	2
002	Protein, Crude (%)	0612	34.44	2.380	20.12	0.2814	0.1478	165	50.87	002.06	2
003	Fat, Crude (%)	0357	1.825	0.1500	3.313	0.2941	0.1237	140	-5.06	003.14	0
003	Fat, Crude (%)	0574	2.260	0.2800	3.313	0.2941	0.1237	140	-3.58	003.14	0
003	Fat, Crude (%)	0265	2.275	0.3700	3.313	0.2941	0.1237	140	-3.53	003.14	0
003	Fat, Crude (%)	2359	2.420	0.3200	3.313	0.2941	0.1237	140	-3.04	003.14	0
003	Fat, Crude (%)	2146	2.460	0.1400	3.313	0.2941	0.1237	140	-2.90	003.14	0
003	Fat, Crude (%)	0885	2.590	0.0200	3.313	0.2941	0.1237	140	-2.46	003.11	0
003	Fat, Crude (%)	0546	2.610	0.0400	3.313	0.2941	0.1237	140	-2.39	003.99	0
003	Fat, Crude (%)	0504	2.645	0.5300	3.313	0.2941	0.1237	140	-2.27	003.01	0
003	Fat, Crude (%)	0144	2.705	0.1100	3.313	0.2941	0.1237	140	-2.07	003.14	0
003	Fat, Crude (%)	0353	2.710	0.3000	3.313	0.2941	0.1237	140	-2.05	003.09	0
003	Fat, Crude (%)	0619	2.810	0.1000	3.313	0.2941	0.1237	140	-1.71	003.10	0
003	Fat, Crude (%)	0229	2.815	0.5100	3.313	0.2941	0.1237	140	-1.69	003.06	0
003	Fat, Crude (%)	0123	2.835	0.0300	3.313	0.2941	0.1237	140	-1.62	003.13	0
003	Fat, Crude (%)	0505	2.875	0.0500	3.313	0.2941	0.1237	140	-1.49	003.14	0
003	Fat, Crude (%)	2129	2.877	0.0230	3.313	0.2941	0.1237	140	-1.48	003.14	0
003	Fat, Crude (%)	0675	2.890	0.0400	3.313	0.2941	0.1237	140	-1.44	003.14	0
003	Fat, Crude (%)	0511	2.910	0.1200	3.313	0.2941	0.1237	140	-1.37	003.14	0
003	Fat, Crude (%)	0683	2.925	0.1100	3.313	0.2941	0.1237	140	-1.32	003.14	0
003	Fat, Crude (%)	2022	2.945	0.0500	3.313	0.2941	0.1237	140	-1.25	003.10	0
003	Fat, Crude (%)	0300	2.965	0.1300	3.313	0.2941	0.1237	140	-1.18	003.14	0
003	Fat, Crude (%)	0083	2.975	0.0100	3.313	0.2941	0.1237	140	-1.15	003.14	0
003	Fat, Crude (%)	2073	3.000	0.0000	3.313	0.2941	0.1237	140	-1.06	003.10	0
003	Fat, Crude (%)	0358	3.010	0.0400	3.313	0.2941	0.1237	140	-1.03	003.14	0
003	Fat, Crude (%)	0037	3.020	0.0800	3.313	0.2941	0.1237	140	-0.99	003.14	0
003	Fat, Crude (%)	0529	3.025	0.0100	3.313	0.2941	0.1237	140	-0.98	003.14	0
003	Fat, Crude (%)	0019	3.025	0.0700	3.313	0.2941	0.1237	140	-0.98	003.14	0
003	Fat, Crude (%)	0848	3.040	0.0200	3.313	0.2941	0.1237	140	-0.93	003.14	0
003	Fat, Crude (%)	0939	3.045	0.0100	3.313	0.2941	0.1237	140	-0.91	003.14	0
003	Fat, Crude (%)	0229	3.050	0.3800	3.313	0.2941	0.1237	140	-0.89	003.14	0
003	Fat, Crude (%)	2089	3.055	0.2100	3.313	0.2941	0.1237	140	-0.88	003.13	0
003	Fat, Crude (%)	0034	3.090	0.2800	3.313	0.2941	0.1237	140	-0.76	003.14	0
003	Fat, Crude (%)	0876	3.100	0.2000	3.313	0.2941	0.1237	140	-0.72	003.14	0
003	Fat, Crude (%)	0956	3.100	0.2000	3.313	0.2941	0.1237	140	-0.72	003.14	0
003	Fat, Crude (%)	0089	3.105	0.0700	3.313	0.2941	0.1237	140	-0.71	003.14	0
003	Fat, Crude (%)	0062	3.120	0.1800	3.313	0.2941	0.1237	140	-0.65	003.10	0
003	Fat, Crude (%)	0407	3.123	0.0638	3.313	0.2941	0.1237	140	-0.64	003.14	0
003	Fat, Crude (%)	0066	3.130	0.0400	3.313	0.2941	0.1237	140	-0.62	003.14	0
003	Fat, Crude (%)	0108	3.130	0.0600	3.313	0.2941	0.1237	140	-0.62	003.14	0
003	Fat, Crude (%)	2196	3.150	0.0000	3.313	0.2941	0.1237	140	-0.55	003.10	0
003	Fat, Crude (%)	2184	3.155	0.3700	3.313	0.2941	0.1237	140	-0.54	003.10	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
003	Fat, Crude (%)	0042	3.190	0.0400	3.313	0.2941	0.1237	140	-0.42	003.10	0
003	Fat, Crude (%)	0171	3.190	0.0600	3.313	0.2941	0.1237	140	-0.42	003.14	0
003	Fat, Crude (%)	2337	3.195	0.0300	3.313	0.2941	0.1237	140	-0.40	003.00	0
003	Fat, Crude (%)	0692	3.200	0.2000	3.313	0.2941	0.1237	140	-0.38	003.14	0
003	Fat, Crude (%)	0853	3.200	0.2000	3.313	0.2941	0.1237	140	-0.38	003.14	0
003	Fat, Crude (%)	2193	3.205	0.0900	3.313	0.2941	0.1237	140	-0.37	003.10	0
003	Fat, Crude (%)	0297	3.220	0.1000	3.313	0.2941	0.1237	140	-0.31	003.06	0
003	Fat, Crude (%)	0298	3.220	0.0200	3.313	0.2941	0.1237	140	-0.31	003.14	0
003	Fat, Crude (%)	0045	3.230	0.2000	3.313	0.2941	0.1237	140	-0.28	003.10	0
003	Fat, Crude (%)	0035	3.230	0.0600	3.313	0.2941	0.1237	140	-0.28	003.14	0
003	Fat, Crude (%)	0074	3.240	0.1400	3.313	0.2941	0.1237	140	-0.25	003.14	0
003	Fat, Crude (%)	0366	3.250	0.1000	3.313	0.2941	0.1237	140	-0.21	003.10	0
003	Fat, Crude (%)	0581	3.250	0.1000	3.313	0.2941	0.1237	140	-0.21	003.14	0
003	Fat, Crude (%)	0009	3.255	0.1100	3.313	0.2941	0.1237	140	-0.20	003.14	0
003	Fat, Crude (%)	0263	3.264	0.0430	3.313	0.2941	0.1237	140	-0.17	003.10	0
003	Fat, Crude (%)	0036	3.276	0.0380	3.313	0.2941	0.1237	140	-0.12	003.10	0
003	Fat, Crude (%)	0889	3.285	0.0300	3.313	0.2941	0.1237	140	-0.09	003.10	0
003	Fat, Crude (%)	0013	3.285	0.2100	3.313	0.2941	0.1237	140	-0.09	003.14	0
003	Fat, Crude (%)	0948	3.290	0.0600	3.313	0.2941	0.1237	140	-0.08	003.09	0
003	Fat, Crude (%)	0001	3.291	0.2648	3.313	0.2941	0.1237	140	-0.07	003.14	0
003	Fat, Crude (%)	0098	3.295	0.0500	3.313	0.2941	0.1237	140	-0.06	003.10	0
003	Fat, Crude (%)	2192	3.300	0.1600	3.313	0.2941	0.1237	140	-0.04	003.00	0
003	Fat, Crude (%)	0689	3.300	0.0000	3.313	0.2941	0.1237	140	-0.04	003.06	0
003	Fat, Crude (%)	0049	3.300	0.2200	3.313	0.2941	0.1237	140	-0.04	003.14	0
003	Fat, Crude (%)	0598	3.305	0.0100	3.313	0.2941	0.1237	140	-0.03	003.14	0
003	Fat, Crude (%)	0870	3.314	0.1514	3.313	0.2941	0.1237	140	0.00	003.10	0
003	Fat, Crude (%)	2076	3.315	0.1973	3.313	0.2941	0.1237	140	0.01	003.14	0
003	Fat, Crude (%)	0693	3.317	0.1000	3.313	0.2941	0.1237	140	0.02	003.10	0
003	Fat, Crude (%)	0033	3.330	0.0200	3.313	0.2941	0.1237	140	0.06	003.13	0
003	Fat, Crude (%)	0098	3.340	0.0000	3.313	0.2941	0.1237	140	0.09	003.13	0
003	Fat, Crude (%)	2155	3.341	0.0890	3.313	0.2941	0.1237	140	0.10	003.10	0
003	Fat, Crude (%)	0644	3.346	0.0080	3.313	0.2941	0.1237	140	0.11	003.06	0
003	Fat, Crude (%)	0354	3.349	0.0030	3.313	0.2941	0.1237	140	0.12	003.09	0
003	Fat, Crude (%)	0683	3.350	0.0600	3.313	0.2941	0.1237	140	0.13	003.00	0
003	Fat, Crude (%)	0689	3.350	0.1000	3.313	0.2941	0.1237	140	0.13	003.10	0
003	Fat, Crude (%)	0278	3.350	0.1000	3.313	0.2941	0.1237	140	0.13	003.14	0
003	Fat, Crude (%)	0884	3.350	0.5000	3.313	0.2941	0.1237	140	0.13	003.06	0
003	Fat, Crude (%)	0918	3.350	0.0000	3.313	0.2941	0.1237	140	0.13	003.06	0
003	Fat, Crude (%)	0880	3.350	0.5000	3.313	0.2941	0.1237	140	0.13	003.99	0
003	Fat, Crude (%)	2340	3.361	0.0020	3.313	0.2941	0.1237	140	0.16	003.00	0
003	Fat, Crude (%)	0948	3.365	0.0100	3.313	0.2941	0.1237	140	0.18	003.14	0
003	Fat, Crude (%)	2325	3.385	0.0500	3.313	0.2941	0.1237	140	0.25	003.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
003	Fat, Crude (%)	0148	3.385	0.2500	3.313	0.2941	0.1237	140	0.25	003.14	0
003	Fat, Crude (%)	2190	3.415	0.2300	3.313	0.2941	0.1237	140	0.35	003.00	0
003	Fat, Crude (%)	0013	3.415	0.0700	3.313	0.2941	0.1237	140	0.35	003.09	0
003	Fat, Crude (%)	0590	3.415	0.0700	3.313	0.2941	0.1237	140	0.35	003.09	0
003	Fat, Crude (%)	0541	3.420	0.0600	3.313	0.2941	0.1237	140	0.37	003.10	0
003	Fat, Crude (%)	0028	3.420	0.1800	3.313	0.2941	0.1237	140	0.37	003.13	0
003	Fat, Crude (%)	0226	3.430	0.0200	3.313	0.2941	0.1237	140	0.40	003.14	0
003	Fat, Crude (%)	0035	3.430	0.0800	3.313	0.2941	0.1237	140	0.40	003.00	0
003	Fat, Crude (%)	0027	3.431	0.0280	3.313	0.2941	0.1237	140	0.40	003.09	0
003	Fat, Crude (%)	2302	3.445	0.1300	3.313	0.2941	0.1237	140	0.45	003.14	0
003	Fat, Crude (%)	0100	3.445	0.0700	3.313	0.2941	0.1237	140	0.45	003.10	0
003	Fat, Crude (%)	2199	3.445	0.4100	3.313	0.2941	0.1237	140	0.45	003.12	0
003	Fat, Crude (%)	2246	3.445	0.0300	3.313	0.2941	0.1237	140	0.45	003.13	0
003	Fat, Crude (%)	0726	3.445	0.0500	3.313	0.2941	0.1237	140	0.45	003.14	0
003	Fat, Crude (%)	0164	3.450	0.0000	3.313	0.2941	0.1237	140	0.47	003.01	0
003	Fat, Crude (%)	0171	3.450	0.1400	3.313	0.2941	0.1237	140	0.47	003.12	0
003	Fat, Crude (%)	2006	3.460	0.0600	3.313	0.2941	0.1237	140	0.50	003.13	0
003	Fat, Crude (%)	0881	3.467	0.1090	3.313	0.2941	0.1237	140	0.52	003.10	0
003	Fat, Crude (%)	0910	3.480	0.0800	3.313	0.2941	0.1237	140	0.57	003.06	0
003	Fat, Crude (%)	0148	3.480	0.0400	3.313	0.2941	0.1237	140	0.57	003.10	0
003	Fat, Crude (%)	2109	3.500	0.1800	3.313	0.2941	0.1237	140	0.64	003.00	0
003	Fat, Crude (%)	0425	3.500	0.0000	3.313	0.2941	0.1237	140	0.64	003.06	0
003	Fat, Crude (%)	0687	3.500	0.2000	3.313	0.2941	0.1237	140	0.64	003.06	0
003	Fat, Crude (%)	0619	3.505	0.0700	3.313	0.2941	0.1237	140	0.65	003.14	0
003	Fat, Crude (%)	0345	3.525	0.1500	3.313	0.2941	0.1237	140	0.72	003.00	0
003	Fat, Crude (%)	0563	3.528	0.0524	3.313	0.2941	0.1237	140	0.73	003.01	0
003	Fat, Crude (%)	0563	3.528	0.0524	3.313	0.2941	0.1237	140	0.73	003.14	0
003	Fat, Crude (%)	0098	3.535	0.1700	3.313	0.2941	0.1237	140	0.76	003.09	0
003	Fat, Crude (%)	0885	3.539	0.0344	3.313	0.2941	0.1237	140	0.77	003.00	0
003	Fat, Crude (%)	2359	3.540	0.2000	3.313	0.2941	0.1237	140	0.77	003.00	0
003	Fat, Crude (%)	0512	3.545	0.2040	3.313	0.2941	0.1237	140	0.79	003.09	0
003	Fat, Crude (%)	0968	3.545	0.3700	3.313	0.2941	0.1237	140	0.79	003.14	0
003	Fat, Crude (%)	0017	3.550	0.1000	3.313	0.2941	0.1237	140	0.81	003.00	0
003	Fat, Crude (%)	2319	3.550	0.1000	3.313	0.2941	0.1237	140	0.81	003.10	0
003	Fat, Crude (%)	0004	3.550	0.5000	3.313	0.2941	0.1237	140	0.81	003.14	0
003	Fat, Crude (%)	0190	3.570	0.0400	3.313	0.2941	0.1237	140	0.88	003.14	0
003	Fat, Crude (%)	0004	3.570	0.0200	3.313	0.2941	0.1237	140	0.88	003.09	0
003	Fat, Crude (%)	0682	3.580	0.0000	3.313	0.2941	0.1237	140	0.91	003.06	0
003	Fat, Crude (%)	0685	3.590	0.1200	3.313	0.2941	0.1237	140	0.94	003.09	0
003	Fat, Crude (%)	0870	3.626	0.0213	3.313	0.2941	0.1237	140	1.07	003.13	0
003	Fat, Crude (%)	0941	3.635	0.0900	3.313	0.2941	0.1237	140	1.10	003.99	0
003	Fat, Crude (%)	0653	3.645	0.0700	3.313	0.2941	0.1237	140	1.13	003.14	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
003	Fat, Crude (%)	0536	3.650	0.3000	3.313	0.2941	0.1237	140	1.15	003.99	0
003	Fat, Crude (%)	2181	3.655	0.2500	3.313	0.2941	0.1237	140	1.16	003.06	0
003	Fat, Crude (%)	0169	3.675	0.1100	3.313	0.2941	0.1237	140	1.23	003.06	0
003	Fat, Crude (%)	0003	3.690	0.3800	3.313	0.2941	0.1237	140	1.28	003.14	0
003	Fat, Crude (%)	0629	3.700	0.0600	3.313	0.2941	0.1237	140	1.32	003.10	0
003	Fat, Crude (%)	0553	3.705	0.0100	3.313	0.2941	0.1237	140	1.33	003.11	0
003	Fat, Crude (%)	0889	3.710	0.0000	3.313	0.2941	0.1237	140	1.35	003.11	0
003	Fat, Crude (%)	0682	3.720	0.0000	3.313	0.2941	0.1237	140	1.39	003.14	0
003	Fat, Crude (%)	0011	3.765	0.0100	3.313	0.2941	0.1237	140	1.54	003.11	0
003	Fat, Crude (%)	0510	3.765	0.0700	3.313	0.2941	0.1237	140	1.54	003.14	0
003	Fat, Crude (%)	0674	3.780	0.0000	3.313	0.2941	0.1237	140	1.59	003.06	0
003	Fat, Crude (%)	0520	3.880	0.3200	3.313	0.2941	0.1237	140	1.93	003.14	0
003	Fat, Crude (%)	0852	3.940	0.0400	3.313	0.2941	0.1237	140	2.13	003.11	0
003	Fat, Crude (%)	2188	4.040	0.4800	3.313	0.2941	0.1237	140	2.47	003.09	0
003	Fat, Crude (%)	2200	4.350	0.3000	3.313	0.2941	0.1237	140	3.53	003.12	0
003	Fat, Crude (%)	0309	4.500	0.0000	3.313	0.2941	0.1237	140	4.04	003.00	0
003	Fat, Crude (%)	2105	10.56	0.0600	3.313	0.2941	0.1237	140	24.65	003.11	2
004	Fiber, Crude (%)	0265	2.105	0.0500	3.465	0.3564	0.1297	109	-3.82	004.07	0
004	Fiber, Crude (%)	2199	2.250	0.1000	3.465	0.3564	0.1297	109	-3.41	004.06	0
004	Fiber, Crude (%)	2188	2.260	0.0000	3.465	0.3564	0.1297	109	-3.38	004.06	0
004	Fiber, Crude (%)	0510	2.450	0.1000	3.465	0.3564	0.1297	109	-2.85	004.07	0
004	Fiber, Crude (%)	0629	2.800	0.0000	3.465	0.3564	0.1297	109	-1.87	004.99	0
004	Fiber, Crude (%)	0708	2.805	0.0300	3.465	0.3564	0.1297	109	-1.85	004.07	0
004	Fiber, Crude (%)	0889	2.820	0.0200	3.465	0.3564	0.1297	109	-1.81	004.11	0
004	Fiber, Crude (%)	0144	2.840	0.2800	3.465	0.3564	0.1297	109	-1.75	004.07	0
004	Fiber, Crude (%)	0278	2.900	0.0000	3.465	0.3564	0.1297	109	-1.59	004.07	0
004	Fiber, Crude (%)	2319	2.950	0.1000	3.465	0.3564	0.1297	109	-1.45	004.06	0
004	Fiber, Crude (%)	2290	2.950	0.1000	3.465	0.3564	0.1297	109	-1.45	004.11	0
004	Fiber, Crude (%)	2193	2.970	0.2600	3.465	0.3564	0.1297	109	-1.39	004.07	0
004	Fiber, Crude (%)	2337	3.010	0.1200	3.465	0.3564	0.1297	109	-1.28	004.07	0
004	Fiber, Crude (%)	2089	3.045	0.1300	3.465	0.3564	0.1297	109	-1.18	004.03	0
004	Fiber, Crude (%)	2009	3.055	0.0374	3.465	0.3564	0.1297	109	-1.15	004.07	0
004	Fiber, Crude (%)	0881	3.065	0.2310	3.465	0.3564	0.1297	109	-1.12	004.07	0
004	Fiber, Crude (%)	2190	3.100	0.0200	3.465	0.3564	0.1297	109	-1.02	004.07	0
004	Fiber, Crude (%)	0028	3.100	0.0000	3.465	0.3564	0.1297	109	-1.02	004.07	0
004	Fiber, Crude (%)	2144	3.125	0.0700	3.465	0.3564	0.1297	109	-0.95	004.01	0
004	Fiber, Crude (%)	0226	3.150	0.1600	3.465	0.3564	0.1297	109	-0.88	004.07	0
004	Fiber, Crude (%)	0169	3.150	0.1000	3.465	0.3564	0.1297	109	-0.88	004.00	0
004	Fiber, Crude (%)	2076	3.186	0.2020	3.465	0.3564	0.1297	109	-0.78	004.00	0
004	Fiber, Crude (%)	0726	3.190	0.0000	3.465	0.3564	0.1297	109	-0.77	004.07	0
004	Fiber, Crude (%)	0425	3.200	0.0000	3.465	0.3564	0.1297	109	-0.74	004.00	0
004	Fiber, Crude (%)	0353	3.200	0.4000	3.465	0.3564	0.1297	109	-0.74	004.03	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
004	Fiber, Crude (%)	0956	3.200	0.0000	3.465	0.3564	0.1297	109	-0.74	004.07	0
004	Fiber, Crude (%)	2340	3.202	0.0260	3.465	0.3564	0.1297	109	-0.74	004.07	0
004	Fiber, Crude (%)	0644	3.218	0.2910	3.465	0.3564	0.1297	109	-0.70	004.07	0
004	Fiber, Crude (%)	0918	3.240	0.1000	3.465	0.3564	0.1297	109	-0.63	004.06	0
004	Fiber, Crude (%)	0619	3.265	0.0500	3.465	0.3564	0.1297	109	-0.56	004.07	0
004	Fiber, Crude (%)	0123	3.285	0.0300	3.465	0.3564	0.1297	109	-0.51	004.07	0
004	Fiber, Crude (%)	0876	3.300	0.4000	3.465	0.3564	0.1297	109	-0.46	004.07	0
004	Fiber, Crude (%)	0553	3.300	0.0000	3.465	0.3564	0.1297	109	-0.46	004.11	0
004	Fiber, Crude (%)	2259	3.323	0.0430	3.465	0.3564	0.1297	109	-0.40	004.06	0
004	Fiber, Crude (%)	0009	3.325	0.1500	3.465	0.3564	0.1297	109	-0.39	004.07	0
004	Fiber, Crude (%)	0300	3.325	0.4300	3.465	0.3564	0.1297	109	-0.39	004.07	0
004	Fiber, Crude (%)	0505	3.335	0.0900	3.465	0.3564	0.1297	109	-0.37	004.07	0
004	Fiber, Crude (%)	0190	3.350	0.1000	3.465	0.3564	0.1297	109	-0.32	004.07	0
004	Fiber, Crude (%)	0581	3.350	0.1000	3.465	0.3564	0.1297	109	-0.32	004.07	0
004	Fiber, Crude (%)	0914	3.350	0.1000	3.465	0.3564	0.1297	109	-0.32	004.07	0
004	Fiber, Crude (%)	0229	3.355	0.1300	3.465	0.3564	0.1297	109	-0.31	004.07	0
004	Fiber, Crude (%)	0263	3.361	0.0220	3.465	0.3564	0.1297	109	-0.29	004.07	0
004	Fiber, Crude (%)	0035	3.375	0.2500	3.465	0.3564	0.1297	109	-0.25	004.07	0
004	Fiber, Crude (%)	0674	3.380	0.0000	3.465	0.3564	0.1297	109	-0.24	004.06	0
004	Fiber, Crude (%)	0366	3.400	0.4000	3.465	0.3564	0.1297	109	-0.18	004.07	0
004	Fiber, Crude (%)	2325	3.410	0.0600	3.465	0.3564	0.1297	109	-0.16	004.06	0
004	Fiber, Crude (%)	0164	3.450	0.1000	3.465	0.3564	0.1297	109	-0.04	004.00	0
004	Fiber, Crude (%)	0083	3.450	0.0600	3.465	0.3564	0.1297	109	-0.04	004.07	0
004	Fiber, Crude (%)	0870	3.452	0.0292	3.465	0.3564	0.1297	109	-0.04	004.07	0
004	Fiber, Crude (%)	0098	3.455	0.0900	3.465	0.3564	0.1297	109	-0.03	004.06	0
004	Fiber, Crude (%)	0881	3.457	0.1160	3.465	0.3564	0.1297	109	-0.02	004.06	0
004	Fiber, Crude (%)	0653	3.465	0.0500	3.465	0.3564	0.1297	109	0.00	004.07	0
004	Fiber, Crude (%)	0675	3.475	0.2300	3.465	0.3564	0.1297	109	0.03	004.07	0
004	Fiber, Crude (%)	0038	3.480	0.1200	3.465	0.3564	0.1297	109	0.04	004.06	0
004	Fiber, Crude (%)	0098	3.485	0.1300	3.465	0.3564	0.1297	109	0.06	004.07	0
004	Fiber, Crude (%)	0948	3.490	0.0800	3.465	0.3564	0.1297	109	0.07	004.07	0
004	Fiber, Crude (%)	0042	3.495	0.4300	3.465	0.3564	0.1297	109	0.08	004.07	0
004	Fiber, Crude (%)	0100	3.500	0.0200	3.465	0.3564	0.1297	109	0.10	004.07	0
004	Fiber, Crude (%)	0407	3.524	0.2494	3.465	0.3564	0.1297	109	0.16	004.07	0
004	Fiber, Crude (%)	0354	3.530	0.0400	3.465	0.3564	0.1297	109	0.18	004.07	0
004	Fiber, Crude (%)	0968	3.530	0.0200	3.465	0.3564	0.1297	109	0.18	004.07	0
004	Fiber, Crude (%)	2302	3.535	0.1900	3.465	0.3564	0.1297	109	0.20	004.07	0
004	Fiber, Crude (%)	0643	3.545	0.0500	3.465	0.3564	0.1297	109	0.22	004.07	0
004	Fiber, Crude (%)	0689	3.550	0.5000	3.465	0.3564	0.1297	109	0.24	004.06	0
004	Fiber, Crude (%)	0683	3.550	0.1000	3.465	0.3564	0.1297	109	0.24	004.07	0
004	Fiber, Crude (%)	0011	3.565	0.0700	3.465	0.3564	0.1297	109	0.28	004.11	0
004	Fiber, Crude (%)	0034	3.570	0.0000	3.465	0.3564	0.1297	109	0.29	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
004	Fiber, Crude (%)	0345	3.575	0.0500	3.465	0.3564	0.1297	109	0.31	004.00	0
004	Fiber, Crude (%)	0968	3.575	0.1100	3.465	0.3564	0.1297	109	0.31	004.06	0
004	Fiber, Crude (%)	0003	3.585	0.0500	3.465	0.3564	0.1297	109	0.34	004.07	0
004	Fiber, Crude (%)	0598	3.595	0.0100	3.465	0.3564	0.1297	109	0.36	004.07	0
004	Fiber, Crude (%)	0013	3.600	0.0800	3.465	0.3564	0.1297	109	0.38	004.00	0
004	Fiber, Crude (%)	0298	3.600	0.2000	3.465	0.3564	0.1297	109	0.38	004.07	0
004	Fiber, Crude (%)	0682	3.600	0.0000	3.465	0.3564	0.1297	109	0.38	004.07	0
004	Fiber, Crude (%)	0939	3.620	0.1800	3.465	0.3564	0.1297	109	0.43	004.07	0
004	Fiber, Crude (%)	0171	3.625	0.0700	3.465	0.3564	0.1297	109	0.45	004.00	0
004	Fiber, Crude (%)	0074	3.635	0.0900	3.465	0.3564	0.1297	109	0.48	004.07	0
004	Fiber, Crude (%)	0309	3.650	0.3000	3.465	0.3564	0.1297	109	0.52	004.00	0
004	Fiber, Crude (%)	0066	3.650	0.1000	3.465	0.3564	0.1297	109	0.52	004.07	0
004	Fiber, Crude (%)	0019	3.655	0.0700	3.465	0.3564	0.1297	109	0.53	004.07	0
004	Fiber, Crude (%)	2129	3.665	0.0100	3.465	0.3564	0.1297	109	0.56	004.99	0
004	Fiber, Crude (%)	2006	3.675	0.4700	3.465	0.3564	0.1297	109	0.59	004.06	0
004	Fiber, Crude (%)	0027	3.692	0.0650	3.465	0.3564	0.1297	109	0.63	004.06	0
004	Fiber, Crude (%)	0692	3.700	0.2000	3.465	0.3564	0.1297	109	0.66	004.07	0
004	Fiber, Crude (%)	0563	3.706	0.1649	3.465	0.3564	0.1297	109	0.68	004.07	0
004	Fiber, Crude (%)	0504	3.710	0.0200	3.465	0.3564	0.1297	109	0.69	004.00	0
004	Fiber, Crude (%)	2181	3.720	0.0600	3.465	0.3564	0.1297	109	0.71	004.07	0
004	Fiber, Crude (%)	0885	3.730	0.0200	3.465	0.3564	0.1297	109	0.74	004.11	0
004	Fiber, Crude (%)	0884	3.750	0.3000	3.465	0.3564	0.1297	109	0.80	004.00	0
004	Fiber, Crude (%)	0910	3.750	0.1000	3.465	0.3564	0.1297	109	0.80	004.07	0
004	Fiber, Crude (%)	0948	3.755	0.0100	3.465	0.3564	0.1297	109	0.81	004.06	0
004	Fiber, Crude (%)	0045	3.770	0.0200	3.465	0.3564	0.1297	109	0.85	004.07	0
004	Fiber, Crude (%)	2196	3.780	0.0000	3.465	0.3564	0.1297	109	0.88	004.07	0
004	Fiber, Crude (%)	0013	3.800	0.1400	3.465	0.3564	0.1297	109	0.94	004.07	0
004	Fiber, Crude (%)	0553	3.825	0.0300	3.465	0.3564	0.1297	109	1.01	004.07	0
004	Fiber, Crude (%)	0529	3.845	0.1100	3.465	0.3564	0.1297	109	1.07	004.07	0
004	Fiber, Crude (%)	0043	3.908	0.4060	3.465	0.3564	0.1297	109	1.24	004.07	0
004	Fiber, Crude (%)	0004	3.975	0.0700	3.465	0.3564	0.1297	109	1.43	004.07	0
004	Fiber, Crude (%)	0626	3.975	0.0700	3.465	0.3564	0.1297	109	1.43	004.07	0
004	Fiber, Crude (%)	0693	3.997	0.5680	3.465	0.3564	0.1297	109	1.49	004.07	0
004	Fiber, Crude (%)	0520	4.150	0.0200	3.465	0.3564	0.1297	109	1.92	004.07	0
004	Fiber, Crude (%)	2022	4.175	0.0100	3.465	0.3564	0.1297	109	1.99	004.06	0
004	Fiber, Crude (%)	2109	4.205	0.4100	3.465	0.3564	0.1297	109	2.08	004.07	0
004	Fiber, Crude (%)	0848	4.235	0.0100	3.465	0.3564	0.1297	109	2.16	004.07	0
004	Fiber, Crude (%)	0852	4.245	0.0900	3.465	0.3564	0.1297	109	2.19	004.11	0
004	Fiber, Crude (%)	2192	4.380	0.4600	3.465	0.3564	0.1297	109	2.57	004.03	0
004	Fiber, Crude (%)	0646	4.380	0.0400	3.465	0.3564	0.1297	109	2.57	004.07	0
004	Fiber, Crude (%)	0511	4.450	0.7000	3.465	0.3564	0.1297	109	2.76	004.07	0
004	Fiber, Crude (%)	0536	4.650	0.3000	3.465	0.3564	0.1297	109	3.32	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
004	Fiber, Crude (%)	2200	2.540	0.9200	3.465	0.3564	0.1297	109	-2.60	004.06	1
004	Fiber, Crude (%)	2246	4.330	0.8400	3.465	0.3564	0.1297	109	2.43	004.03	1
005	Ash (%)	0309	3.590	0.0200	6.621	0.2420	0.0747	128	-12.52	005.00	0
005	Ash (%)	0511	6.100	0.0000	6.621	0.2420	0.0747	128	-2.15	005.00	0
005	Ash (%)	2184	6.170	0.2800	6.621	0.2420	0.0747	128	-1.86	005.00	0
005	Ash (%)	0019	6.200	0.0000	6.621	0.2420	0.0747	128	-1.74	005.00	0
005	Ash (%)	0169	6.215	0.0700	6.621	0.2420	0.0747	128	-1.68	005.00	0
005	Ash (%)	2268	6.230	0.0400	6.621	0.2420	0.0747	128	-1.62	005.00	0
005	Ash (%)	0853	6.250	0.3000	6.621	0.2420	0.0747	128	-1.53	005.00	0
005	Ash (%)	0536	6.250	0.1000	6.621	0.2420	0.0747	128	-1.53	005.99	0
005	Ash (%)	0953	6.265	0.0700	6.621	0.2420	0.0747	128	-1.47	005.00	0
005	Ash (%)	2188	6.280	0.0000	6.621	0.2420	0.0747	128	-1.41	005.00	0
005	Ash (%)	0358	6.295	0.0500	6.621	0.2420	0.0747	128	-1.35	005.00	0
005	Ash (%)	0004	6.300	0.0600	6.621	0.2420	0.0747	128	-1.33	005.00	0
005	Ash (%)	0539	6.305	0.1300	6.621	0.2420	0.0747	128	-1.31	005.00	0
005	Ash (%)	0885	6.305	0.0300	6.621	0.2420	0.0747	128	-1.31	005.11	0
005	Ash (%)	0049	6.330	0.0800	6.621	0.2420	0.0747	128	-1.20	005.00	0
005	Ash (%)	0357	6.350	0.1000	6.621	0.2420	0.0747	128	-1.12	005.00	0
005	Ash (%)	2200	6.350	0.1000	6.621	0.2420	0.0747	128	-1.12	005.00	0
005	Ash (%)	0918	6.355	0.0700	6.621	0.2420	0.0747	128	-1.10	005.00	0
005	Ash (%)	0726	6.355	0.0100	6.621	0.2420	0.0747	128	-1.10	005.05	0
005	Ash (%)	0644	6.363	0.0190	6.621	0.2420	0.0747	128	-1.07	005.05	0
005	Ash (%)	0278	6.365	0.2300	6.621	0.2420	0.0747	128	-1.06	005.00	0
005	Ash (%)	0881	6.387	0.0420	6.621	0.2420	0.0747	128	-0.97	005.00	0
005	Ash (%)	0651	6.392	0.0077	6.621	0.2420	0.0747	128	-0.94	005.00	0
005	Ash (%)	0229	6.395	0.0900	6.621	0.2420	0.0747	128	-0.93	005.00	0
005	Ash (%)	0038	6.405	0.0900	6.621	0.2420	0.0747	128	-0.89	005.05	0
005	Ash (%)	0650	6.420	0.1600	6.621	0.2420	0.0747	128	-0.83	005.00	0
005	Ash (%)	0529	6.425	0.1500	6.621	0.2420	0.0747	128	-0.81	005.00	0
005	Ash (%)	0066	6.445	0.0300	6.621	0.2420	0.0747	128	-0.73	005.05	0
005	Ash (%)	0001	6.447	0.0941	6.621	0.2420	0.0747	128	-0.72	005.00	0
005	Ash (%)	0646	6.450	0.0200	6.621	0.2420	0.0747	128	-0.71	005.00	0
005	Ash (%)	2199	6.450	0.1000	6.621	0.2420	0.0747	128	-0.71	005.00	0
005	Ash (%)	0042	6.465	0.0100	6.621	0.2420	0.0747	128	-0.64	005.00	0
005	Ash (%)	0027	6.468	0.0450	6.621	0.2420	0.0747	128	-0.63	005.00	0
005	Ash (%)	0682	6.470	0.0000	6.621	0.2420	0.0747	128	-0.62	005.00	0
005	Ash (%)	0574	6.475	0.0900	6.621	0.2420	0.0747	128	-0.60	005.00	0
005	Ash (%)	0164	6.480	0.2200	6.621	0.2420	0.0747	128	-0.58	005.00	0
005	Ash (%)	0692	6.485	0.0500	6.621	0.2420	0.0747	128	-0.56	005.05	0
005	Ash (%)	0003	6.490	0.0200	6.621	0.2420	0.0747	128	-0.54	005.05	0
005	Ash (%)	0693	6.493	0.3050	6.621	0.2420	0.0747	128	-0.53	005.00	0
005	Ash (%)	0598	6.495	0.0100	6.621	0.2420	0.0747	128	-0.52	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
005	Ash (%)	0366	6.500	0.0000	6.621	0.2420	0.0747	128	-0.50	005.00	0
005	Ash (%)	2290	6.500	0.0000	6.621	0.2420	0.0747	128	-0.50	005.11	0
005	Ash (%)	2340	6.504	0.0050	6.621	0.2420	0.0747	128	-0.48	005.00	0
005	Ash (%)	0674	6.510	0.0000	6.621	0.2420	0.0747	128	-0.46	005.00	0
005	Ash (%)	0541	6.520	0.1800	6.621	0.2420	0.0747	128	-0.42	005.00	0
005	Ash (%)	2022	6.520	0.0800	6.621	0.2420	0.0747	128	-0.42	005.00	0
005	Ash (%)	2337	6.520	0.0200	6.621	0.2420	0.0747	128	-0.42	005.00	0
005	Ash (%)	0520	6.545	0.0700	6.621	0.2420	0.0747	128	-0.31	005.00	0
005	Ash (%)	0083	6.550	0.0800	6.621	0.2420	0.0747	128	-0.29	005.00	0
005	Ash (%)	0956	6.550	0.1000	6.621	0.2420	0.0747	128	-0.29	005.00	0
005	Ash (%)	0144	6.560	0.1000	6.621	0.2420	0.0747	128	-0.25	005.00	0
005	Ash (%)	0353	6.565	0.0300	6.621	0.2420	0.0747	128	-0.23	005.00	0
005	Ash (%)	0148	6.570	0.1800	6.621	0.2420	0.0747	128	-0.21	005.00	0
005	Ash (%)	2150	6.570	0.0200	6.621	0.2420	0.0747	128	-0.21	005.00	0
005	Ash (%)	0186	6.586	0.1060	6.621	0.2420	0.0747	128	-0.14	005.00	0
005	Ash (%)	0870	6.587	0.0069	6.621	0.2420	0.0747	128	-0.14	005.00	0
005	Ash (%)	0553	6.590	0.1400	6.621	0.2420	0.0747	128	-0.13	005.00	0
005	Ash (%)	0880	6.590	0.2200	6.621	0.2420	0.0747	128	-0.13	005.00	0
005	Ash (%)	0263	6.590	0.0140	6.621	0.2420	0.0747	128	-0.13	005.05	0
005	Ash (%)	0876	6.600	0.2000	6.621	0.2420	0.0747	128	-0.09	005.00	0
005	Ash (%)	0546	6.615	0.0900	6.621	0.2420	0.0747	128	-0.02	005.99	0
005	Ash (%)	0968	6.620	0.0200	6.621	0.2420	0.0747	128	0.00	005.00	0
005	Ash (%)	2006	6.625	0.0900	6.621	0.2420	0.0747	128	0.02	005.00	0
005	Ash (%)	0407	6.630	0.0890	6.621	0.2420	0.0747	128	0.04	005.00	0
005	Ash (%)	0062	6.630	0.1600	6.621	0.2420	0.0747	128	0.04	005.00	0
005	Ash (%)	0425	6.630	0.2000	6.621	0.2420	0.0747	128	0.04	005.00	0
005	Ash (%)	0563	6.633	0.1398	6.621	0.2420	0.0747	128	0.05	005.00	0
005	Ash (%)	0300	6.635	0.1300	6.621	0.2420	0.0747	128	0.06	005.00	0
005	Ash (%)	2196	6.640	0.0000	6.621	0.2420	0.0747	128	0.08	005.00	0
005	Ash (%)	0123	6.645	0.0500	6.621	0.2420	0.0747	128	0.10	005.00	0
005	Ash (%)	2302	6.655	0.2300	6.621	0.2420	0.0747	128	0.14	005.00	0
005	Ash (%)	2246	6.660	0.0000	6.621	0.2420	0.0747	128	0.16	005.00	0
005	Ash (%)	2009	6.666	0.0514	6.621	0.2420	0.0747	128	0.18	005.05	0
005	Ash (%)	0100	6.670	0.1400	6.621	0.2420	0.0747	128	0.20	005.00	0
005	Ash (%)	0298	6.670	0.0200	6.621	0.2420	0.0747	128	0.20	005.00	0
005	Ash (%)	0643	6.675	0.0100	6.621	0.2420	0.0747	128	0.22	005.00	0
005	Ash (%)	0098	6.695	0.0700	6.621	0.2420	0.0747	128	0.31	005.00	0
005	Ash (%)	0142	6.700	0.0400	6.621	0.2420	0.0747	128	0.33	005.00	0
005	Ash (%)	0265	6.700	0.0000	6.621	0.2420	0.0747	128	0.33	005.00	0
005	Ash (%)	2190	6.700	0.1000	6.621	0.2420	0.0747	128	0.33	005.05	0
005	Ash (%)	0884	6.700	0.0000	6.621	0.2420	0.0747	128	0.33	005.99	0
005	Ash (%)	0910	6.700	0.0000	6.621	0.2420	0.0747	128	0.33	005.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
005	Ash (%)	2076	6.711	0.0055	6.621	0.2420	0.0747	128	0.37	005.00	0
005	Ash (%)	0354	6.720	0.0200	6.621	0.2420	0.0747	128	0.41	005.00	0
005	Ash (%)	2105	6.720	0.1200	6.621	0.2420	0.0747	128	0.41	005.11	0
005	Ash (%)	0689	6.730	0.0400	6.621	0.2420	0.0747	128	0.45	005.05	0
005	Ash (%)	0035	6.735	0.0100	6.621	0.2420	0.0747	128	0.47	005.00	0
005	Ash (%)	0047	6.735	0.1700	6.621	0.2420	0.0747	128	0.47	005.00	0
005	Ash (%)	0510	6.745	0.0900	6.621	0.2420	0.0747	128	0.51	005.00	0
005	Ash (%)	2109	6.750	0.1000	6.621	0.2420	0.0747	128	0.53	005.05	0
005	Ash (%)	2319	6.750	0.1000	6.621	0.2420	0.0747	128	0.53	005.99	0
005	Ash (%)	0881	6.757	0.0960	6.621	0.2420	0.0747	128	0.56	005.05	0
005	Ash (%)	0889	6.760	0.0200	6.621	0.2420	0.0747	128	0.58	005.05	0
005	Ash (%)	2259	6.767	0.0110	6.621	0.2420	0.0747	128	0.60	005.00	0
005	Ash (%)	0683	6.770	0.1000	6.621	0.2420	0.0747	128	0.62	005.05	0
005	Ash (%)	0619	6.775	0.0700	6.621	0.2420	0.0747	128	0.64	005.05	0
005	Ash (%)	2193	6.780	0.1000	6.621	0.2420	0.0747	128	0.66	005.00	0
005	Ash (%)	0653	6.785	0.0100	6.621	0.2420	0.0747	128	0.68	005.05	0
005	Ash (%)	0017	6.800	0.0000	6.621	0.2420	0.0747	128	0.74	005.00	0
005	Ash (%)	2181	6.805	0.0900	6.621	0.2420	0.0747	128	0.76	005.00	0
005	Ash (%)	0848	6.810	0.0000	6.621	0.2420	0.0747	128	0.78	005.00	0
005	Ash (%)	2146	6.815	0.0500	6.621	0.2420	0.0747	128	0.80	005.05	0
005	Ash (%)	0675	6.825	0.1100	6.621	0.2420	0.0747	128	0.84	005.00	0
005	Ash (%)	0590	6.830	0.0200	6.621	0.2420	0.0747	128	0.86	005.05	0
005	Ash (%)	2192	6.845	0.1300	6.621	0.2420	0.0747	128	0.93	005.05	0
005	Ash (%)	0885	6.846	0.0684	6.621	0.2420	0.0747	128	0.93	005.05	0
005	Ash (%)	0948	6.855	0.0300	6.621	0.2420	0.0747	128	0.97	005.00	0
005	Ash (%)	0345	6.860	0.0400	6.621	0.2420	0.0747	128	0.99	005.00	0
005	Ash (%)	0629	6.860	0.0000	6.621	0.2420	0.0747	128	0.99	005.99	0
005	Ash (%)	0504	6.865	0.1500	6.621	0.2420	0.0747	128	1.01	005.00	0
005	Ash (%)	0644	6.870	0.0170	6.621	0.2420	0.0747	128	1.03	005.02	0
005	Ash (%)	2325	6.880	0.0200	6.621	0.2420	0.0747	128	1.07	005.00	0
005	Ash (%)	0939	6.880	0.0400	6.621	0.2420	0.0747	128	1.07	005.05	0
005	Ash (%)	2246	6.925	0.0100	6.621	0.2420	0.0747	128	1.26	005.05	0
005	Ash (%)	2129	6.930	0.0400	6.621	0.2420	0.0747	128	1.28	005.05	0
005	Ash (%)	0941	6.945	0.1300	6.621	0.2420	0.0747	128	1.34	005.05	0
005	Ash (%)	0034	6.946	0.0230	6.621	0.2420	0.0747	128	1.34	005.05	0
005	Ash (%)	0581	6.950	0.1000	6.621	0.2420	0.0747	128	1.36	005.00	0
005	Ash (%)	0171	6.960	0.1400	6.621	0.2420	0.0747	128	1.40	005.00	0
005	Ash (%)	2089	6.970	0.1200	6.621	0.2420	0.0747	128	1.44	005.00	0
005	Ash (%)	0297	6.995	0.1100	6.621	0.2420	0.0747	128	1.55	005.00	0
005	Ash (%)	0652	7.000	0.0000	6.621	0.2420	0.0747	128	1.57	005.99	0
005	Ash (%)	0190	7.135	0.0100	6.621	0.2420	0.0747	128	2.12	005.05	0
005	Ash (%)	0685	7.180	0.1000	6.621	0.2420	0.0747	128	2.31	005.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
005	Ash (%)	0226	7.395	0.1700	6.621	0.2420	0.0747	128	3.20	005.00	0
005	Ash (%)	0889	8.920	0.0600	6.621	0.2420	0.0747	128	9.50	005.11	0
005	Ash (%)	0852	9.755	0.0300	6.621	0.2420	0.0747	128	12.95	005.11	0
005	Ash (%)	0914	10.00	0.2200	6.621	0.2420	0.0747	128	13.96	005.00	0
005	Ash (%)	0108	5.725	0.7900	6.621	0.2420	0.0747	128	-3.70	005.00	1
005	Ash (%)	0045	6.485	0.4100	6.621	0.2420	0.0747	128	-0.56	005.00	1
005	Ash (%)	0852	0.9350	0.0100	6.621	0.2420	0.0747	128	-23.50	005.00	2
006	Total Sugars (%)	2290	3.200	0.0000	4.216	0.7609	0.2675	11	-1.34	006.99	0
006	Total Sugars (%)	0939	3.395	1.050	4.216	0.7609	0.2675	11	-1.08	006.99	0
006	Total Sugars (%)	0941	3.730	0.3600	4.216	0.7609	0.2675	11	-0.64	006.99	0
006	Total Sugars (%)	0885	4.021	0.2596	4.216	0.7609	0.2675	11	-0.26	006.00	0
006	Total Sugars (%)	0910	4.135	0.5500	4.216	0.7609	0.2675	11	-0.11	006.00	0
006	Total Sugars (%)	2192	4.170	0.0800	4.216	0.7609	0.2675	11	-0.06	006.00	0
006	Total Sugars (%)	2319	4.300	0.0000	4.216	0.7609	0.2675	11	0.11	006.00	0
006	Total Sugars (%)	0921	4.342	0.0230	4.216	0.7609	0.2675	11	0.16	006.00	0
006	Total Sugars (%)	0956	4.450	0.5000	4.216	0.7609	0.2675	11	0.31	006.99	0
006	Total Sugars (%)	0226	5.280	0.0800	4.216	0.7609	0.2675	11	1.40	006.99	0
006	Total Sugars (%)	0407	5.440	0.0400	4.216	0.7609	0.2675	11	1.61	006.00	0
008	Fiber, Acid Detergent (%)	0889	3.130	0.1000	5.131	0.5861	0.1648	60	-3.41	008.08	0
008	Fiber, Acid Detergent (%)	0848	4.105	0.0500	5.131	0.5861	0.1648	60	-1.75	008.08	0
008	Fiber, Acid Detergent (%)	2290	4.200	0.0000	5.131	0.5861	0.1648	60	-1.59	008.99	0
008	Fiber, Acid Detergent (%)	0939	4.320	0.0600	5.131	0.5861	0.1648	60	-1.38	008.08	0
008	Fiber, Acid Detergent (%)	2302	4.345	0.2300	5.131	0.5861	0.1648	60	-1.34	008.08	0
008	Fiber, Acid Detergent (%)	0675	4.390	0.0800	5.131	0.5861	0.1648	60	-1.26	008.08	0
008	Fiber, Acid Detergent (%)	0956	4.400	0.0000	5.131	0.5861	0.1648	60	-1.25	008.08	0
008	Fiber, Acid Detergent (%)	0098	4.540	0.4600	5.131	0.5861	0.1648	60	-1.01	008.02	0
008	Fiber, Acid Detergent (%)	0278	4.550	0.5000	5.131	0.5861	0.1648	60	-0.99	008.08	0
008	Fiber, Acid Detergent (%)	0674	4.600	0.0000	5.131	0.5861	0.1648	60	-0.91	008.02	0
008	Fiber, Acid Detergent (%)	0590	4.620	0.4800	5.131	0.5861	0.1648	60	-0.87	008.08	0
008	Fiber, Acid Detergent (%)	0581	4.700	0.2000	5.131	0.5861	0.1648	60	-0.74	008.08	0
008	Fiber, Acid Detergent (%)	0013	4.710	0.3200	5.131	0.5861	0.1648	60	-0.72	008.08	0
008	Fiber, Acid Detergent (%)	0653	4.730	0.1400	5.131	0.5861	0.1648	60	-0.68	008.08	0
008	Fiber, Acid Detergent (%)	0968	4.730	0.0600	5.131	0.5861	0.1648	60	-0.68	008.08	0
008	Fiber, Acid Detergent (%)	2129	4.741	0.0240	5.131	0.5861	0.1648	60	-0.67	008.08	0
008	Fiber, Acid Detergent (%)	0004	4.750	0.2600	5.131	0.5861	0.1648	60	-0.65	008.08	0
008	Fiber, Acid Detergent (%)	0910	4.750	0.0400	5.131	0.5861	0.1648	60	-0.65	008.08	0
008	Fiber, Acid Detergent (%)	0504	4.760	0.1200	5.131	0.5861	0.1648	60	-0.63	008.02	0
008	Fiber, Acid Detergent (%)	0629	4.850	0.1000	5.131	0.5861	0.1648	60	-0.48	008.99	0
008	Fiber, Acid Detergent (%)	0353	4.940	0.1800	5.131	0.5861	0.1648	60	-0.33	008.02	0
008	Fiber, Acid Detergent (%)	0309	4.950	0.1000	5.131	0.5861	0.1648	60	-0.31	008.02	0
008	Fiber, Acid Detergent (%)	0263	4.968	0.0490	5.131	0.5861	0.1648	60	-0.28	008.08	0
008	Fiber, Acid Detergent (%)	0357	4.970	0.3000	5.131	0.5861	0.1648	60	-0.28	008.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
008	Fiber, Acid Detergent (%)	2192	4.975	0.1100	5.131	0.5861	0.1648	60	-0.27	008.02	0
008	Fiber, Acid Detergent (%)	0016	4.985	0.1500	5.131	0.5861	0.1648	60	-0.25	008.08	0
008	Fiber, Acid Detergent (%)	0226	4.990	0.2000	5.131	0.5861	0.1648	60	-0.24	008.02	0
008	Fiber, Acid Detergent (%)	0297	5.000	0.4000	5.131	0.5861	0.1648	60	-0.22	008.08	0
008	Fiber, Acid Detergent (%)	2190	5.005	0.0100	5.131	0.5861	0.1648	60	-0.22	008.08	0
008	Fiber, Acid Detergent (%)	0619	5.075	0.3100	5.131	0.5861	0.1648	60	-0.10	008.08	0
008	Fiber, Acid Detergent (%)	0027	5.084	0.0070	5.131	0.5861	0.1648	60	-0.08	008.99	0
008	Fiber, Acid Detergent (%)	0190	5.110	0.0600	5.131	0.5861	0.1648	60	-0.04	008.08	0
008	Fiber, Acid Detergent (%)	0726	5.110	0.0600	5.131	0.5861	0.1648	60	-0.04	008.08	0
008	Fiber, Acid Detergent (%)	0354	5.150	0.1400	5.131	0.5861	0.1648	60	0.03	008.08	0
008	Fiber, Acid Detergent (%)	0345	5.250	0.1000	5.131	0.5861	0.1648	60	0.20	008.02	0
008	Fiber, Acid Detergent (%)	0148	5.270	0.1400	5.131	0.5861	0.1648	60	0.24	008.02	0
008	Fiber, Acid Detergent (%)	0512	5.284	0.1900	5.131	0.5861	0.1648	60	0.26	008.02	0
008	Fiber, Acid Detergent (%)	0693	5.300	0.2560	5.131	0.5861	0.1648	60	0.29	008.08	0
008	Fiber, Acid Detergent (%)	0148	5.310	0.1200	5.131	0.5861	0.1648	60	0.31	008.08	0
008	Fiber, Acid Detergent (%)	0358	5.335	0.0100	5.131	0.5861	0.1648	60	0.35	008.08	0
008	Fiber, Acid Detergent (%)	0042	5.355	0.5300	5.131	0.5861	0.1648	60	0.38	008.08	0
008	Fiber, Acid Detergent (%)	0870	5.398	0.0287	5.131	0.5861	0.1648	60	0.46	008.08	0
008	Fiber, Acid Detergent (%)	0948	5.435	0.4700	5.131	0.5861	0.1648	60	0.52	008.02	0
008	Fiber, Acid Detergent (%)	0001	5.481	0.1580	5.131	0.5861	0.1648	60	0.60	008.08	0
008	Fiber, Acid Detergent (%)	0066	5.485	0.2900	5.131	0.5861	0.1648	60	0.60	008.08	0
008	Fiber, Acid Detergent (%)	2196	5.590	0.0000	5.131	0.5861	0.1648	60	0.78	008.08	0
008	Fiber, Acid Detergent (%)	0918	5.595	0.1500	5.131	0.5861	0.1648	60	0.79	008.08	0
008	Fiber, Acid Detergent (%)	2193	5.665	0.0300	5.131	0.5861	0.1648	60	0.91	008.08	0
008	Fiber, Acid Detergent (%)	0083	5.675	0.0900	5.131	0.5861	0.1648	60	0.93	008.08	0
008	Fiber, Acid Detergent (%)	0035	5.680	0.1400	5.131	0.5861	0.1648	60	0.94	008.08	0
008	Fiber, Acid Detergent (%)	0689	5.750	0.3000	5.131	0.5861	0.1648	60	1.06	008.02	0
008	Fiber, Acid Detergent (%)	0037	5.780	0.0000	5.131	0.5861	0.1648	60	1.11	008.08	0
008	Fiber, Acid Detergent (%)	0510	5.800	0.4000	5.131	0.5861	0.1648	60	1.14	008.08	0
008	Fiber, Acid Detergent (%)	0045	5.855	0.1100	5.131	0.5861	0.1648	60	1.23	008.08	0
008	Fiber, Acid Detergent (%)	0646	5.975	0.1100	5.131	0.5861	0.1648	60	1.44	008.08	0
008	Fiber, Acid Detergent (%)	0407	6.060	0.1561	5.131	0.5861	0.1648	60	1.59	008.08	0
008	Fiber, Acid Detergent (%)	0164	6.100	0.0000	5.131	0.5861	0.1648	60	1.65	008.08	0
008	Fiber, Acid Detergent (%)	0511	6.450	0.5000	5.131	0.5861	0.1648	60	2.25	008.08	0
008	Fiber, Acid Detergent (%)	0265	6.985	0.1100	5.131	0.5861	0.1648	60	3.16	008.05	0
008	Fiber, Acid Detergent (%)	0536	10.10	0.2000	5.131	0.5861	0.1648	60	8.48	008.08	0
008	Fiber, Acid Detergent (%)	0049	4.780	0.8000	5.131	0.5861	0.1648	60	-0.60	008.08	1
008	Fiber, Acid Detergent (%)	2246	16.85	0.2024	5.131	0.5861	0.1648	60	20.00	008.02	2
009	Fiber, Neutral Detergent (%)	0889	9.355	0.0700	13.16	1.109	0.3409	55	-3.43	009.99	0
009	Fiber, Neutral Detergent (%)	0889	10.82	0.3700	13.16	1.109	0.3409	55	-2.11	009.09	0
009	Fiber, Neutral Detergent (%)	0848	11.16	0.7100	13.16	1.109	0.3409	55	-1.80	009.09	0
009	Fiber, Neutral Detergent (%)	0598	11.26	0.0600	13.16	1.109	0.3409	55	-1.71	009.09	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
009	Fiber, Neutral Detergent (%)	0910	11.80	0.0000	13.16	1.109	0.3409	55	-1.22	009.07	0
009	Fiber, Neutral Detergent (%)	0098	11.90	0.2000	13.16	1.109	0.3409	55	-1.13	009.07	0
009	Fiber, Neutral Detergent (%)	0939	12.09	1.230	13.16	1.109	0.3409	55	-0.97	009.09	0
009	Fiber, Neutral Detergent (%)	0510	12.20	0.4000	13.16	1.109	0.3409	55	-0.86	009.09	0
009	Fiber, Neutral Detergent (%)	0512	12.26	0.2900	13.16	1.109	0.3409	55	-0.81	009.07	0
009	Fiber, Neutral Detergent (%)	0870	12.26	0.3375	13.16	1.109	0.3409	55	-0.81	009.09	0
009	Fiber, Neutral Detergent (%)	0581	12.30	0.2000	13.16	1.109	0.3409	55	-0.77	009.09	0
009	Fiber, Neutral Detergent (%)	0357	12.32	0.0200	13.16	1.109	0.3409	55	-0.75	009.09	0
009	Fiber, Neutral Detergent (%)	0309	12.37	0.6477	13.16	1.109	0.3409	55	-0.70	009.07	0
009	Fiber, Neutral Detergent (%)	0263	12.40	0.2510	13.16	1.109	0.3409	55	-0.68	009.09	0
009	Fiber, Neutral Detergent (%)	2340	12.48	0.0130	13.16	1.109	0.3409	55	-0.60	009.09	0
009	Fiber, Neutral Detergent (%)	0646	12.51	0.0400	13.16	1.109	0.3409	55	-0.58	009.09	0
009	Fiber, Neutral Detergent (%)	0190	12.55	0.6600	13.16	1.109	0.3409	55	-0.55	009.09	0
009	Fiber, Neutral Detergent (%)	0726	12.55	0.6600	13.16	1.109	0.3409	55	-0.55	009.09	0
009	Fiber, Neutral Detergent (%)	0353	12.57	0.2100	13.16	1.109	0.3409	55	-0.53	009.07	0
009	Fiber, Neutral Detergent (%)	0619	12.60	0.3200	13.16	1.109	0.3409	55	-0.50	009.09	0
009	Fiber, Neutral Detergent (%)	2190	12.63	0.1400	13.16	1.109	0.3409	55	-0.47	009.09	0
009	Fiber, Neutral Detergent (%)	0629	12.75	1.100	13.16	1.109	0.3409	55	-0.37	009.99	0
009	Fiber, Neutral Detergent (%)	0956	12.85	0.1000	13.16	1.109	0.3409	55	-0.28	009.09	0
009	Fiber, Neutral Detergent (%)	0083	12.90	0.0100	13.16	1.109	0.3409	55	-0.23	009.09	0
009	Fiber, Neutral Detergent (%)	0358	12.92	0.7400	13.16	1.109	0.3409	55	-0.21	009.09	0
009	Fiber, Neutral Detergent (%)	0045	13.00	1.000	13.16	1.109	0.3409	55	-0.14	009.09	0
009	Fiber, Neutral Detergent (%)	0148	13.05	0.1800	13.16	1.109	0.3409	55	-0.10	009.09	0
009	Fiber, Neutral Detergent (%)	0504	13.06	0.6200	13.16	1.109	0.3409	55	-0.09	009.04	0
009	Fiber, Neutral Detergent (%)	0675	13.13	0.1900	13.16	1.109	0.3409	55	-0.03	009.09	0
009	Fiber, Neutral Detergent (%)	2302	13.18	0.2000	13.16	1.109	0.3409	55	0.02	009.09	0
009	Fiber, Neutral Detergent (%)	2290	13.20	0.4000	13.16	1.109	0.3409	55	0.04	009.99	0
009	Fiber, Neutral Detergent (%)	0693	13.20	0.5260	13.16	1.109	0.3409	55	0.04	009.09	0
009	Fiber, Neutral Detergent (%)	0354	13.23	0.4200	13.16	1.109	0.3409	55	0.07	009.09	0
009	Fiber, Neutral Detergent (%)	0948	13.32	0.1900	13.16	1.109	0.3409	55	0.14	009.07	0
009	Fiber, Neutral Detergent (%)	2193	13.40	0.0100	13.16	1.109	0.3409	55	0.22	009.09	0
009	Fiber, Neutral Detergent (%)	0265	13.50	0.2000	13.16	1.109	0.3409	55	0.31	009.09	0
009	Fiber, Neutral Detergent (%)	0066	13.65	0.2500	13.16	1.109	0.3409	55	0.44	009.09	0
009	Fiber, Neutral Detergent (%)	0674	13.70	0.0000	13.16	1.109	0.3409	55	0.49	009.07	0
009	Fiber, Neutral Detergent (%)	0037	13.71	0.4800	13.16	1.109	0.3409	55	0.50	009.09	0
009	Fiber, Neutral Detergent (%)	2196	13.76	0.0000	13.16	1.109	0.3409	55	0.55	009.09	0
009	Fiber, Neutral Detergent (%)	0278	13.80	0.2000	13.16	1.109	0.3409	55	0.58	009.09	0
009	Fiber, Neutral Detergent (%)	2192	13.86	0.2000	13.16	1.109	0.3409	55	0.64	009.07	0
009	Fiber, Neutral Detergent (%)	0918	13.87	0.4800	13.16	1.109	0.3409	55	0.64	009.09	0
009	Fiber, Neutral Detergent (%)	0226	14.00	0.2900	13.16	1.109	0.3409	55	0.76	009.07	0
009	Fiber, Neutral Detergent (%)	0653	14.04	0.3100	13.16	1.109	0.3409	55	0.79	009.09	0
009	Fiber, Neutral Detergent (%)	0689	14.10	0.6000	13.16	1.109	0.3409	55	0.85	009.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
009	Fiber, Neutral Detergent (%)	2129	14.48	0.4900	13.16	1.109	0.3409	55	1.19	009.09	0
009	Fiber, Neutral Detergent (%)	0035	14.51	0.7400	13.16	1.109	0.3409	55	1.22	009.09	0
009	Fiber, Neutral Detergent (%)	2246	14.75	0.0769	13.16	1.109	0.3409	55	1.43	009.07	0
009	Fiber, Neutral Detergent (%)	0968	14.98	0.2400	13.16	1.109	0.3409	55	1.65	009.09	0
009	Fiber, Neutral Detergent (%)	0407	15.06	0.4748	13.16	1.109	0.3409	55	1.72	009.09	0
009	Fiber, Neutral Detergent (%)	0590	15.45	0.2000	13.16	1.109	0.3409	55	2.07	009.09	0
009	Fiber, Neutral Detergent (%)	0164	16.25	0.3000	13.16	1.109	0.3409	55	2.79	009.09	0
009	Fiber, Neutral Detergent (%)	0345	17.50	0.6000	13.16	1.109	0.3409	55	3.92	009.07	0
009	Fiber, Neutral Detergent (%)	0536	18.55	0.1000	13.16	1.109	0.3409	55	4.86	009.09	0
009	Fiber, Neutral Detergent (%)	0049	13.09	1.970	13.16	1.109	0.3409	55	-0.06	009.09	1
009	Fiber, Neutral Detergent (%)	0297	19.90	2.600	13.16	1.109	0.3409	55	6.08	009.09	1
009	Fiber, Neutral Detergent (%)	2246	27.04	0.1300	13.16	1.109	0.3409	55	12.52	009.04	2
010	Moisture (%)	2329	8.295	0.3300	9.184	0.5249	0.0677	22	-1.69	010.99	0
010	Moisture (%)	0914	8.380	0.2200	9.184	0.5249	0.0677	22	-1.53	010.99	0
010	Moisture (%)	0546	8.545	0.0500	9.184	0.5249	0.0677	22	-1.22	010.99	0
010	Moisture (%)	2129	8.635	0.0100	9.184	0.5249	0.0677	22	-1.05	010.99	0
010	Moisture (%)	0889	8.695	0.0300	9.184	0.5249	0.0677	22	-0.93	010.11	0
010	Moisture (%)	0843	8.925	0.0700	9.184	0.5249	0.0677	22	-0.49	010.03	0
010	Moisture (%)	2089	8.960	0.0400	9.184	0.5249	0.0677	22	-0.43	010.99	0
010	Moisture (%)	0939	9.030	0.1200	9.184	0.5249	0.0677	22	-0.29	010.99	0
010	Moisture (%)	0948	9.080	0.0200	9.184	0.5249	0.0677	22	-0.20	010.99	0
010	Moisture (%)	0726	9.100	0.0000	9.184	0.5249	0.0677	22	-0.16	010.99	0
010	Moisture (%)	0164	9.240	0.0000	9.184	0.5249	0.0677	22	0.11	010.03	0
010	Moisture (%)	2290	9.285	0.1900	9.184	0.5249	0.0677	22	0.19	010.11	0
010	Moisture (%)	2076	9.290	0.2200	9.184	0.5249	0.0677	22	0.20	010.99	0
010	Moisture (%)	0652	9.300	0.0000	9.184	0.5249	0.0677	22	0.22	010.99	0
010	Moisture (%)	0921	9.350	0.0200	9.184	0.5249	0.0677	22	0.32	010.99	0
010	Moisture (%)	2150	9.420	0.0200	9.184	0.5249	0.0677	22	0.45	010.99	0
010	Moisture (%)	0884	9.500	0.0000	9.184	0.5249	0.0677	22	0.60	010.99	0
010	Moisture (%)	0553	9.535	0.0300	9.184	0.5249	0.0677	22	0.67	010.11	0
010	Moisture (%)	0190	9.615	0.0300	9.184	0.5249	0.0677	22	0.82	010.99	0
010	Moisture (%)	0885	9.805	0.0500	9.184	0.5249	0.0677	22	1.18	010.11	0
010	Moisture (%)	2105	10.24	0.0000	9.184	0.5249	0.0677	22	2.01	010.11	0
010	Moisture (%)	0852	11.18	0.0400	9.184	0.5249	0.0677	22	3.80	010.11	0
010	Moisture (%)	0692	9.165	0.8500	9.184	0.5249	0.0677	22	-0.04	010.99	1
011	Loss on Drying, HT (%)	0685	8.720	0.0800	9.904	0.3570	0.1129	68	-3.32	011.01	0
011	Loss on Drying, HT (%)	2188	8.900	0.0000	9.904	0.3570	0.1129	68	-2.81	011.01	0
011	Loss on Drying, HT (%)	0859	9.037	0.0520	9.904	0.3570	0.1129	68	-2.43	011.01	0
011	Loss on Drying, HT (%)	0520	9.170	0.0400	9.904	0.3570	0.1129	68	-2.06	011.01	0
011	Loss on Drying, HT (%)	2168	9.240	0.2600	9.904	0.3570	0.1129	68	-1.86	011.99	0
011	Loss on Drying, HT (%)	2200	9.400	0.2000	9.904	0.3570	0.1129	68	-1.41	011.01	0
011	Loss on Drying, HT (%)	2340	9.435	0.0220	9.904	0.3570	0.1129	68	-1.31	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
011	Loss on Drying, HT (%)	0881	9.452	0.1140	9.904	0.3570	0.1129	68	-1.27	011.01	0
011	Loss on Drying, HT (%)	2006	9.485	0.4100	9.904	0.3570	0.1129	68	-1.17	011.01	0
011	Loss on Drying, HT (%)	2181	9.535	0.2300	9.904	0.3570	0.1129	68	-1.03	011.01	0
011	Loss on Drying, HT (%)	2359	9.550	0.3000	9.904	0.3570	0.1129	68	-0.99	011.01	0
011	Loss on Drying, HT (%)	0968	9.560	0.0200	9.904	0.3570	0.1129	68	-0.96	011.01	0
011	Loss on Drying, HT (%)	2303	9.567	0.1840	9.904	0.3570	0.1129	68	-0.94	011.99	0
011	Loss on Drying, HT (%)	0539	9.600	0.0200	9.904	0.3570	0.1129	68	-0.85	011.01	0
011	Loss on Drying, HT (%)	0953	9.620	0.0000	9.904	0.3570	0.1129	68	-0.79	011.01	0
011	Loss on Drying, HT (%)	0563	9.675	0.0178	9.904	0.3570	0.1129	68	-0.64	011.01	0
011	Loss on Drying, HT (%)	0425	9.685	0.0900	9.904	0.3570	0.1129	68	-0.61	011.01	0
011	Loss on Drying, HT (%)	2337	9.705	0.0100	9.904	0.3570	0.1129	68	-0.56	011.01	0
011	Loss on Drying, HT (%)	0942	9.710	0.0200	9.904	0.3570	0.1129	68	-0.54	011.02	0
011	Loss on Drying, HT (%)	0674	9.730	0.0000	9.904	0.3570	0.1129	68	-0.49	011.01	0
011	Loss on Drying, HT (%)	2174	9.760	0.4600	9.904	0.3570	0.1129	68	-0.40	011.99	0
011	Loss on Drying, HT (%)	0358	9.775	0.0900	9.904	0.3570	0.1129	68	-0.36	011.01	0
011	Loss on Drying, HT (%)	2196	9.780	0.0000	9.904	0.3570	0.1129	68	-0.35	011.01	0
011	Loss on Drying, HT (%)	0354	9.785	0.0100	9.904	0.3570	0.1129	68	-0.33	011.01	0
011	Loss on Drying, HT (%)	0298	9.790	0.0200	9.904	0.3570	0.1129	68	-0.32	011.01	0
011	Loss on Drying, HT (%)	0510	9.800	0.2000	9.904	0.3570	0.1129	68	-0.29	011.01	0
011	Loss on Drying, HT (%)	0598	9.805	0.0100	9.904	0.3570	0.1129	68	-0.28	011.01	0
011	Loss on Drying, HT (%)	2022	9.815	0.2700	9.904	0.3570	0.1129	68	-0.25	011.01	0
011	Loss on Drying, HT (%)	0100	9.820	0.1200	9.904	0.3570	0.1129	68	-0.23	011.01	0
011	Loss on Drying, HT (%)	0574	9.820	0.0000	9.904	0.3570	0.1129	68	-0.23	011.01	0
011	Loss on Drying, HT (%)	2325	9.830	0.0200	9.904	0.3570	0.1129	68	-0.21	011.01	0
011	Loss on Drying, HT (%)	0852	9.840	0.2400	9.904	0.3570	0.1129	68	-0.18	011.99	0
011	Loss on Drying, HT (%)	0675	9.900	0.0400	9.904	0.3570	0.1129	68	-0.01	011.01	0
011	Loss on Drying, HT (%)	2193	9.905	0.0500	9.904	0.3570	0.1129	68	0.00	011.01	0
011	Loss on Drying, HT (%)	0263	9.910	0.0120	9.904	0.3570	0.1129	68	0.02	011.01	0
011	Loss on Drying, HT (%)	0108	9.920	0.1600	9.904	0.3570	0.1129	68	0.05	011.01	0
011	Loss on Drying, HT (%)	0013	9.950	0.3800	9.904	0.3570	0.1129	68	0.13	011.01	0
011	Loss on Drying, HT (%)	0643	9.955	0.0500	9.904	0.3570	0.1129	68	0.14	011.01	0
011	Loss on Drying, HT (%)	0541	9.960	0.0000	9.904	0.3570	0.1129	68	0.16	011.01	0
011	Loss on Drying, HT (%)	0300	9.995	0.2100	9.904	0.3570	0.1129	68	0.26	011.01	0
011	Loss on Drying, HT (%)	0226	10.00	0.1800	9.904	0.3570	0.1129	68	0.27	011.01	0
011	Loss on Drying, HT (%)	2246	10.01	0.0200	9.904	0.3570	0.1129	68	0.30	011.01	0
011	Loss on Drying, HT (%)	0098	10.02	0.0600	9.904	0.3570	0.1129	68	0.33	011.01	0
011	Loss on Drying, HT (%)	0682	10.02	0.0000	9.904	0.3570	0.1129	68	0.33	011.01	0
011	Loss on Drying, HT (%)	0870	10.02	0.0480	9.904	0.3570	0.1129	68	0.34	011.01	0
011	Loss on Drying, HT (%)	0529	10.04	0.0200	9.904	0.3570	0.1129	68	0.38	011.02	0
011	Loss on Drying, HT (%)	0171	10.06	0.0300	9.904	0.3570	0.1129	68	0.42	011.01	0
011	Loss on Drying, HT (%)	0164	10.06	0.1600	9.904	0.3570	0.1129	68	0.44	011.01	0
011	Loss on Drying, HT (%)	0066	10.07	0.0900	9.904	0.3570	0.1129	68	0.45	011.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
011	Loss on Drying, HT (%)	0653	10.08	0.0300	9.904	0.3570	0.1129	68	0.48	011.01	0
011	Loss on Drying, HT (%)	0553	10.09	0.0300	9.904	0.3570	0.1129	68	0.51	011.01	0
011	Loss on Drying, HT (%)	0309	10.10	0.0200	9.904	0.3570	0.1129	68	0.55	011.01	0
011	Loss on Drying, HT (%)	2302	10.11	0.0400	9.904	0.3570	0.1129	68	0.58	011.01	0
011	Loss on Drying, HT (%)	2268	10.12	0.0500	9.904	0.3570	0.1129	68	0.59	011.01	0
011	Loss on Drying, HT (%)	0511	10.18	0.1000	9.904	0.3570	0.1129	68	0.77	011.01	0
011	Loss on Drying, HT (%)	2184	10.23	0.2100	9.904	0.3570	0.1129	68	0.90	011.01	0
011	Loss on Drying, HT (%)	0880	10.24	0.4400	9.904	0.3570	0.1129	68	0.94	011.01	0
011	Loss on Drying, HT (%)	2073	10.30	0.1900	9.904	0.3570	0.1129	68	1.10	011.01	0
011	Loss on Drying, HT (%)	0144	10.32	0.1500	9.904	0.3570	0.1129	68	1.15	011.01	0
011	Loss on Drying, HT (%)	0123	10.34	0.0400	9.904	0.3570	0.1129	68	1.22	011.01	0
011	Loss on Drying, HT (%)	0265	10.40	0.2000	9.904	0.3570	0.1129	68	1.39	011.01	0
011	Loss on Drying, HT (%)	0047	10.41	0.4100	9.904	0.3570	0.1129	68	1.40	011.01	0
011	Loss on Drying, HT (%)	0062	10.41	0.1800	9.904	0.3570	0.1129	68	1.42	011.01	0
011	Loss on Drying, HT (%)	0004	10.45	0.0600	9.904	0.3570	0.1129	68	1.53	011.01	0
011	Loss on Drying, HT (%)	0229	10.59	0.1900	9.904	0.3570	0.1129	68	1.91	011.01	0
011	Loss on Drying, HT (%)	2168	10.59	0.2100	9.904	0.3570	0.1129	68	1.91	011.01	0
011	Loss on Drying, HT (%)	2058	10.63	0.0300	9.904	0.3570	0.1129	68	2.02	011.01	0
011	Loss on Drying, HT (%)	0407	10.65	0.0793	9.904	0.3570	0.1129	68	2.08	011.01	0
011	Loss on Drying, HT (%)	0651	9.486	0.6694	9.904	0.3570	0.1129	68	-1.17	011.01	1
011	Loss on Drying, HT (%)	0843	11.33	0.7900	9.904	0.3570	0.1129	68	3.98	011.01	1
011	Loss on Drying, HT (%)	0536	6.450	0.1000	9.904	0.3570	0.1129	68	-9.67	011.01	2
012	Starch (%)	0265	31.45	0.5000	39.23	2.132	0.3992	37	-3.65	012.01	0
012	Starch (%)	0889	34.42	0.0900	39.23	2.132	0.3992	37	-2.26	012.11	0
012	Starch (%)	0852	34.86	0.0100	39.23	2.132	0.3992	37	-2.05	012.11	0
012	Starch (%)	0918	35.32	0.4780	39.23	2.132	0.3992	37	-1.84	012.01	0
012	Starch (%)	0910	36.95	1.100	39.23	2.132	0.3992	37	-1.07	012.01	0
012	Starch (%)	0510	37.20	0.6000	39.23	2.132	0.3992	37	-0.95	012.04	0
012	Starch (%)	0870	37.40	0.5734	39.23	2.132	0.3992	37	-0.86	012.01	0
012	Starch (%)	0407	37.40	0.2140	39.23	2.132	0.3992	37	-0.86	012.03	0
012	Starch (%)	2129	37.56	0.6845	39.23	2.132	0.3992	37	-0.78	012.04	0
012	Starch (%)	0407	37.71	0.0657	39.23	2.132	0.3992	37	-0.71	012.01	0
012	Starch (%)	0164	37.80	0.4000	39.23	2.132	0.3992	37	-0.67	012.01	0
012	Starch (%)	0885	38.06	0.1300	39.23	2.132	0.3992	37	-0.55	012.11	0
012	Starch (%)	2181	38.37	0.4400	39.23	2.132	0.3992	37	-0.40	012.00	0
012	Starch (%)	0885	38.65	0.1590	39.23	2.132	0.3992	37	-0.27	012.01	0
012	Starch (%)	0098	38.80	1.800	39.23	2.132	0.3992	37	-0.20	012.04	0
012	Starch (%)	0148	38.91	0.1100	39.23	2.132	0.3992	37	-0.15	012.01	0
012	Starch (%)	2192	38.91	0.6600	39.23	2.132	0.3992	37	-0.15	012.00	0
012	Starch (%)	0689	39.10	0.2000	39.23	2.132	0.3992	37	-0.06	012.00	0
012	Starch (%)	0045	39.35	1.300	39.23	2.132	0.3992	37	0.05	012.01	0
012	Starch (%)	2193	39.44	0.2100	39.23	2.132	0.3992	37	0.09	012.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
012	Starch (%)	0278	39.45	0.1000	39.23	2.132	0.3992	37	0.10	012.04	0
012	Starch (%)	0683	39.98	0.0500	39.23	2.132	0.3992	37	0.35	012.00	0
012	Starch (%)	0939	40.00	0.2200	39.23	2.132	0.3992	37	0.36	012.00	0
012	Starch (%)	0226	40.16	0.3600	39.23	2.132	0.3992	37	0.43	012.04	0
012	Starch (%)	0164	40.40	0.4000	39.23	2.132	0.3992	37	0.55	012.00	0
012	Starch (%)	0354	40.56	0.0500	39.23	2.132	0.3992	37	0.62	012.00	0
012	Starch (%)	0297	40.59	1.060	39.23	2.132	0.3992	37	0.64	012.03	0
012	Starch (%)	0619	40.70	0.2000	39.23	2.132	0.3992	37	0.69	012.00	0
012	Starch (%)	2188	40.81	0.0000	39.23	2.132	0.3992	37	0.74	012.00	0
012	Starch (%)	2006	40.96	0.1700	39.23	2.132	0.3992	37	0.81	012.00	0
012	Starch (%)	2196	41.17	0.0000	39.23	2.132	0.3992	37	0.91	012.00	0
012	Starch (%)	0353	41.51	0.3000	39.23	2.132	0.3992	37	1.07	012.20	0
012	Starch (%)	0693	41.51	0.1140	39.23	2.132	0.3992	37	1.07	012.01	0
012	Starch (%)	0948	41.61	0.9400	39.23	2.132	0.3992	37	1.11	012.04	0
012	Starch (%)	2290	42.00	0.4000	39.23	2.132	0.3992	37	1.30	012.11	0
012	Starch (%)	0910	42.05	0.5000	39.23	2.132	0.3992	37	1.32	012.00	0
012	Starch (%)	2089	43.12	0.1800	39.23	2.132	0.3992	37	1.82	012.01	0
012	Starch (%)	0956	36.50	2.200	39.23	2.132	0.3992	37	-1.28	012.01	1
013	Fat, Acid Pretreat (%)	0651	2.292	0.0222	4.093	0.5487	0.1430	51	-3.28	013.13	0
013	Fat, Acid Pretreat (%)	0885	2.590	0.0200	4.093	0.5487	0.1430	51	-2.74	013.10	0
013	Fat, Acid Pretreat (%)	2129	2.877	0.0230	4.093	0.5487	0.1430	51	-2.22	013.13	0
013	Fat, Acid Pretreat (%)	2073	3.115	0.0300	4.093	0.5487	0.1430	51	-1.78	013.00	0
013	Fat, Acid Pretreat (%)	2290	3.150	0.1000	4.093	0.5487	0.1430	51	-1.72	013.12	0
013	Fat, Acid Pretreat (%)	2268	3.220	0.0400	4.093	0.5487	0.1430	51	-1.59	013.10	0
013	Fat, Acid Pretreat (%)	0036	3.298	0.0513	4.093	0.5487	0.1430	51	-1.45	013.10	0
013	Fat, Acid Pretreat (%)	0675	3.545	0.0500	4.093	0.5487	0.1430	51	-1.00	013.02	0
013	Fat, Acid Pretreat (%)	2129	3.624	0.2780	4.093	0.5487	0.1430	51	-0.86	013.00	0
013	Fat, Acid Pretreat (%)	0083	3.665	0.0900	4.093	0.5487	0.1430	51	-0.78	013.00	0
013	Fat, Acid Pretreat (%)	0683	3.670	0.0200	4.093	0.5487	0.1430	51	-0.77	013.13	0
013	Fat, Acid Pretreat (%)	2146	3.730	0.1400	4.093	0.5487	0.1430	51	-0.66	013.00	0
013	Fat, Acid Pretreat (%)	0027	3.736	0.2350	4.093	0.5487	0.1430	51	-0.65	013.13	0
013	Fat, Acid Pretreat (%)	2196	3.780	0.0000	4.093	0.5487	0.1430	51	-0.57	013.10	0
013	Fat, Acid Pretreat (%)	0297	3.800	0.1400	4.093	0.5487	0.1430	51	-0.53	013.00	0
013	Fat, Acid Pretreat (%)	0939	3.825	0.3100	4.093	0.5487	0.1430	51	-0.49	013.13	0
013	Fat, Acid Pretreat (%)	0689	3.850	0.1000	4.093	0.5487	0.1430	51	-0.44	013.00	0
013	Fat, Acid Pretreat (%)	2144	3.855	0.1300	4.093	0.5487	0.1430	51	-0.43	013.00	0
013	Fat, Acid Pretreat (%)	0650	3.865	0.0500	4.093	0.5487	0.1430	51	-0.42	013.00	0
013	Fat, Acid Pretreat (%)	0948	3.975	0.0500	4.093	0.5487	0.1430	51	-0.22	013.00	0
013	Fat, Acid Pretreat (%)	0353	3.995	0.0300	4.093	0.5487	0.1430	51	-0.18	013.10	0
013	Fat, Acid Pretreat (%)	0148	4.000	0.5200	4.093	0.5487	0.1430	51	-0.17	013.02	0
013	Fat, Acid Pretreat (%)	0843	4.040	0.4800	4.093	0.5487	0.1430	51	-0.10	013.13	0
013	Fat, Acid Pretreat (%)	0581	4.050	0.1000	4.093	0.5487	0.1430	51	-0.08	013.13	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
013	Fat, Acid Pretreat (%)	0098	4.095	0.2700	4.093	0.5487	0.1430	51	0.00	013.02	0
013	Fat, Acid Pretreat (%)	0265	4.175	0.1300	4.093	0.5487	0.1430	51	0.15	013.13	0
013	Fat, Acid Pretreat (%)	2319	4.200	0.2000	4.093	0.5487	0.1430	51	0.19	013.00	0
013	Fat, Acid Pretreat (%)	0910	4.225	0.1300	4.093	0.5487	0.1430	51	0.24	013.00	0
013	Fat, Acid Pretreat (%)	2329	4.265	0.0100	4.093	0.5487	0.1430	51	0.31	013.02	0
013	Fat, Acid Pretreat (%)	0229	4.270	0.4000	4.093	0.5487	0.1430	51	0.32	013.00	0
013	Fat, Acid Pretreat (%)	0541	4.285	0.0500	4.093	0.5487	0.1430	51	0.35	013.00	0
013	Fat, Acid Pretreat (%)	0652	4.300	0.0000	4.093	0.5487	0.1430	51	0.38	013.00	0
013	Fat, Acid Pretreat (%)	0881	4.314	0.0390	4.093	0.5487	0.1430	51	0.40	013.00	0
013	Fat, Acid Pretreat (%)	0968	4.345	0.2500	4.093	0.5487	0.1430	51	0.46	013.13	0
013	Fat, Acid Pretreat (%)	0853	4.400	0.2000	4.093	0.5487	0.1430	51	0.56	013.02	0
013	Fat, Acid Pretreat (%)	0914	4.400	0.0200	4.093	0.5487	0.1430	51	0.56	013.02	0
013	Fat, Acid Pretreat (%)	0643	4.415	0.0300	4.093	0.5487	0.1430	51	0.59	013.02	0
013	Fat, Acid Pretreat (%)	0921	4.415	0.1500	4.093	0.5487	0.1430	51	0.59	013.02	0
013	Fat, Acid Pretreat (%)	0038	4.435	0.1100	4.093	0.5487	0.1430	51	0.62	013.02	0
013	Fat, Acid Pretreat (%)	0682	4.480	0.0000	4.093	0.5487	0.1430	51	0.71	013.02	0
013	Fat, Acid Pretreat (%)	2150	4.490	0.3000	4.093	0.5487	0.1430	51	0.72	013.00	0
013	Fat, Acid Pretreat (%)	0164	4.510	0.0400	4.093	0.5487	0.1430	51	0.76	013.02	0
013	Fat, Acid Pretreat (%)	2259	4.572	0.0530	4.093	0.5487	0.1430	51	0.87	013.02	0
013	Fat, Acid Pretreat (%)	2076	4.583	0.0662	4.093	0.5487	0.1430	51	0.89	013.00	0
013	Fat, Acid Pretreat (%)	0553	4.615	0.2900	4.093	0.5487	0.1430	51	0.95	013.02	0
013	Fat, Acid Pretreat (%)	0918	4.715	0.1900	4.093	0.5487	0.1430	51	1.13	013.10	0
013	Fat, Acid Pretreat (%)	0870	4.785	0.0751	4.093	0.5487	0.1430	51	1.26	013.02	0
013	Fat, Acid Pretreat (%)	0016	4.790	0.0200	4.093	0.5487	0.1430	51	1.27	013.02	0
013	Fat, Acid Pretreat (%)	0171	5.025	0.5300	4.093	0.5487	0.1430	51	1.70	013.02	0
013	Fat, Acid Pretreat (%)	0309	5.455	0.0110	4.093	0.5487	0.1430	51	2.48	013.00	0
013	Fat, Acid Pretreat (%)	0407	5.760	0.7200	4.093	0.5487	0.1430	51	3.04	013.02	0
013	Fat, Acid Pretreat (%)	0407	5.400	0.9661	4.093	0.5487	0.1430	51	2.38	013.13	1
015	Aluminum (ppm)	0037	41.65	1.900	80.50	18.90	2.165	17	-2.06	015.42	0
015	Aluminum (ppm)	0049	52.90	3.120	80.50	18.90	2.165	17	-1.46	015.41	0
015	Aluminum (ppm)	0171	54.43	2.470	80.50	18.90	2.165	17	-1.38	015.41	0
015	Aluminum (ppm)	0425	69.82	1.530	80.50	18.90	2.165	17	-0.57	015.43	0
015	Aluminum (ppm)	0297	75.00	2.000	80.50	18.90	2.165	17	-0.29	015.43	0
015	Aluminum (ppm)	0407	79.41	3.711	80.50	18.90	2.165	17	-0.06	015.41	0
015	Aluminum (ppm)	0169	79.65	0.5000	80.50	18.90	2.165	17	-0.04	015.43	0
015	Aluminum (ppm)	0407	80.02	1.494	80.50	18.90	2.165	17	-0.03	015.43	0
015	Aluminum (ppm)	0939	80.58	1.710	80.50	18.90	2.165	17	0.00	015.53	0
015	Aluminum (ppm)	0510	81.00	6.000	80.50	18.90	2.165	17	0.03	015.43	0
015	Aluminum (ppm)	0407	84.05	1.626	80.50	18.90	2.165	17	0.19	015.53	0
015	Aluminum (ppm)	0553	87.45	0.1000	80.50	18.90	2.165	17	0.37	015.53	0
015	Aluminum (ppm)	0164	88.50	1.000	80.50	18.90	2.165	17	0.42	015.41	0
015	Aluminum (ppm)	0345	90.45	2.500	80.50	18.90	2.165	17	0.53	015.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
015	Aluminum (ppm)	0520	95.40	1.150	80.50	18.90	2.165	17	0.79	015.41	0
015	Aluminum (ppm)	0353	109.5	5.000	80.50	18.90	2.165	17	1.53	015.43	0
015	Aluminum (ppm)	2129	137.2	1.000	80.50	18.90	2.165	17	3.00	015.42	0
017	Boron (ppm)	0265	7.520	0.8800	9.163	0.9727	0.5816	17	-1.69	017.42	0
017	Boron (ppm)	0358	8.290	0.3800	9.163	0.9727	0.5816	17	-0.90	017.41	0
017	Boron (ppm)	0045	8.315	0.3300	9.163	0.9727	0.5816	17	-0.87	017.42	0
017	Boron (ppm)	0229	8.370	1.000	9.163	0.9727	0.5816	17	-0.82	017.41	0
017	Boron (ppm)	0407	8.374	0.8301	9.163	0.9727	0.5816	17	-0.81	017.53	0
017	Boron (ppm)	0693	8.444	0.2290	9.163	0.9727	0.5816	17	-0.74	017.42	0
017	Boron (ppm)	0345	8.600	0.2000	9.163	0.9727	0.5816	17	-0.58	017.43	0
017	Boron (ppm)	0407	9.061	0.0560	9.163	0.9727	0.5816	17	-0.11	017.41	0
017	Boron (ppm)	0037	9.350	1.500	9.163	0.9727	0.5816	17	0.19	017.42	0
017	Boron (ppm)	0407	9.399	1.425	9.163	0.9727	0.5816	17	0.24	017.43	0
017	Boron (ppm)	0226	9.495	0.2300	9.163	0.9727	0.5816	17	0.34	017.41	0
017	Boron (ppm)	0353	9.500	1.000	9.163	0.9727	0.5816	17	0.35	017.43	0
017	Boron (ppm)	2129	9.655	0.0080	9.163	0.9727	0.5816	17	0.51	017.42	0
017	Boron (ppm)	0510	10.00	0.0000	9.163	0.9727	0.5816	17	0.86	017.43	0
017	Boron (ppm)	0049	10.06	0.9300	9.163	0.9727	0.5816	17	0.92	017.41	0
017	Boron (ppm)	0918	10.54	0.7900	9.163	0.9727	0.5816	17	1.41	017.43	0
017	Boron (ppm)	0083	13.55	0.1000	9.163	0.9727	0.5816	17	4.51	017.43	0
017	Boron (ppm)	0297	0.0000	0.0000	9.163	0.9727	0.5816	17		017.43	4
019	Calcium (%)	0852	0.6000	0.2000	1.349	0.0821	0.0364	121	-9.12	019.08	0
019	Calcium (%)	2006	0.7100	0.0160	1.349	0.0821	0.0364	121	-7.78	019.00	0
019	Calcium (%)	2006	0.7330	0.0280	1.349	0.0821	0.0364	121	-7.50	019.09	0
019	Calcium (%)	2129	1.014	0.0020	1.349	0.0821	0.0364	121	-4.08	019.42	0
019	Calcium (%)	0047	1.105	0.1700	1.349	0.0821	0.0364	121	-2.97	019.52	0
019	Calcium (%)	0536	1.150	0.1000	1.349	0.0821	0.0364	121	-2.42	019.31	0
019	Calcium (%)	2325	1.180	0.0000	1.349	0.0821	0.0364	121	-2.06	019.31	0
019	Calcium (%)	0612	1.185	0.0500	1.349	0.0821	0.0364	121	-2.00	019.33	0
019	Calcium (%)	0948	1.190	0.0000	1.349	0.0821	0.0364	121	-1.93	019.00	0
019	Calcium (%)	2319	1.195	0.0500	1.349	0.0821	0.0364	121	-1.87	019.42	0
019	Calcium (%)	0889	1.195	0.0100	1.349	0.0821	0.0364	121	-1.87	019.99	0
019	Calcium (%)	2200	1.225	0.0500	1.349	0.0821	0.0364	121	-1.51	019.08	0
019	Calcium (%)	0939	1.225	0.0700	1.349	0.0821	0.0364	121	-1.51	019.53	0
019	Calcium (%)	2246	1.230	0.0000	1.349	0.0821	0.0364	121	-1.45	019.31	0
019	Calcium (%)	0407	1.235	0.0445	1.349	0.0821	0.0364	121	-1.39	019.53	0
019	Calcium (%)	2259	1.242	0.0160	1.349	0.0821	0.0364	121	-1.30	019.41	0
019	Calcium (%)	0171	1.245	0.0300	1.349	0.0821	0.0364	121	-1.26	019.41	0
019	Calcium (%)	0154	1.248	0.0056	1.349	0.0821	0.0364	121	-1.23	019.52	0
019	Calcium (%)	0300	1.250	0.0000	1.349	0.0821	0.0364	121	-1.20	019.33	0
019	Calcium (%)	2302	1.260	0.0200	1.349	0.0821	0.0364	121	-1.08	019.00	0
019	Calcium (%)	0045	1.275	0.0900	1.349	0.0821	0.0364	121	-0.90	019.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
019	Calcium (%)	0407	1.280	0.0404	1.349	0.0821	0.0364	121	-0.84	019.41	0
019	Calcium (%)	0848	1.280	0.0000	1.349	0.0821	0.0364	121	-0.84	019.41	0
019	Calcium (%)	0651	1.285	0.0428	1.349	0.0821	0.0364	121	-0.78	019.31	0
019	Calcium (%)	0263	1.285	0.0030	1.349	0.0821	0.0364	121	-0.78	019.42	0
019	Calcium (%)	0033	1.285	0.0300	1.349	0.0821	0.0364	121	-0.78	019.43	0
019	Calcium (%)	0009	1.286	0.0023	1.349	0.0821	0.0364	121	-0.77	019.42	0
019	Calcium (%)	0083	1.290	0.0200	1.349	0.0821	0.0364	121	-0.72	019.43	0
019	Calcium (%)	0186	1.294	0.0035	1.349	0.0821	0.0364	121	-0.67	019.52	0
019	Calcium (%)	0123	1.295	0.0100	1.349	0.0821	0.0364	121	-0.66	019.41	0
019	Calcium (%)	2109	1.297	0.0200	1.349	0.0821	0.0364	121	-0.63	019.00	0
019	Calcium (%)	0590	1.300	0.0000	1.349	0.0821	0.0364	121	-0.59	019.08	0
019	Calcium (%)	0572	1.300	0.0200	1.349	0.0821	0.0364	121	-0.59	019.53	0
019	Calcium (%)	0918	1.301	0.0740	1.349	0.0821	0.0364	121	-0.58	019.43	0
019	Calcium (%)	0511	1.305	0.1700	1.349	0.0821	0.0364	121	-0.53	019.43	0
019	Calcium (%)	0019	1.305	0.0300	1.349	0.0821	0.0364	121	-0.53	019.41	0
019	Calcium (%)	0510	1.310	0.0200	1.349	0.0821	0.0364	121	-0.47	019.43	0
019	Calcium (%)	0529	1.315	0.0300	1.349	0.0821	0.0364	121	-0.41	019.31	0
019	Calcium (%)	2009	1.320	0.0035	1.349	0.0821	0.0364	121	-0.35	019.08	0
019	Calcium (%)	2196	1.320	0.0000	1.349	0.0821	0.0364	121	-0.35	019.31	0
019	Calcium (%)	0004	1.320	0.0000	1.349	0.0821	0.0364	121	-0.35	019.41	0
019	Calcium (%)	0590	1.320	0.1000	1.349	0.0821	0.0364	121	-0.35	019.99	0
019	Calcium (%)	2196	1.324	0.0000	1.349	0.0821	0.0364	121	-0.30	019.08	0
019	Calcium (%)	2337	1.325	0.0100	1.349	0.0821	0.0364	121	-0.29	019.08	0
019	Calcium (%)	0884	1.325	0.0100	1.349	0.0821	0.0364	121	-0.29	019.31	0
019	Calcium (%)	0098	1.325	0.0700	1.349	0.0821	0.0364	121	-0.29	019.41	0
019	Calcium (%)	2089	1.325	0.0300	1.349	0.0821	0.0364	121	-0.29	019.43	0
019	Calcium (%)	0870	1.330	0.0910	1.349	0.0821	0.0364	121	-0.23	019.42	0
019	Calcium (%)	0169	1.330	0.0200	1.349	0.0821	0.0364	121	-0.23	019.32	0
019	Calcium (%)	0653	1.337	0.0080	1.349	0.0821	0.0364	121	-0.14	019.41	0
019	Calcium (%)	0298	1.340	0.0200	1.349	0.0821	0.0364	121	-0.11	019.41	0
019	Calcium (%)	0876	1.345	0.0900	1.349	0.0821	0.0364	121	-0.05	019.43	0
019	Calcium (%)	0164	1.345	0.0300	1.349	0.0821	0.0364	121	-0.05	019.44	0
019	Calcium (%)	0017	1.345	0.0103	1.349	0.0821	0.0364	121	-0.04	019.43	0
019	Calcium (%)	0036	1.347	0.0325	1.349	0.0821	0.0364	121	-0.03	019.43	0
019	Calcium (%)	0885	1.349	0.0084	1.349	0.0821	0.0364	121	0.00	019.08	0
019	Calcium (%)	2181	1.350	0.1000	1.349	0.0821	0.0364	121	0.01	019.00	0
019	Calcium (%)	0675	1.350	0.0200	1.349	0.0821	0.0364	121	0.01	019.43	0
019	Calcium (%)	0407	1.350	0.0146	1.349	0.0821	0.0364	121	0.02	019.43	0
019	Calcium (%)	0186	1.352	0.0390	1.349	0.0821	0.0364	121	0.03	019.42	0
019	Calcium (%)	0881	1.353	0.0140	1.349	0.0821	0.0364	121	0.05	019.00	0
019	Calcium (%)	0689	1.355	0.0300	1.349	0.0821	0.0364	121	0.08	019.08	0
019	Calcium (%)	2340	1.356	0.0030	1.349	0.0821	0.0364	121	0.08	019.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
019	Calcium (%)	0043	1.358	0.0191	1.349	0.0821	0.0364	121	0.12	019.43	0
019	Calcium (%)	0553	1.360	0.0400	1.349	0.0821	0.0364	121	0.14	019.53	0
019	Calcium (%)	0629	1.360	0.0000	1.349	0.0821	0.0364	121	0.14	019.08	0
019	Calcium (%)	2022	1.360	0.0000	1.349	0.0821	0.0364	121	0.14	019.31	0
019	Calcium (%)	0674	1.360	0.0000	1.349	0.0821	0.0364	121	0.14	019.41	0
019	Calcium (%)	0278	1.360	0.0200	1.349	0.0821	0.0364	121	0.14	019.42	0
019	Calcium (%)	0108	1.365	0.0300	1.349	0.0821	0.0364	121	0.20	019.31	0
019	Calcium (%)	0265	1.365	0.1700	1.349	0.0821	0.0364	121	0.20	019.42	0
019	Calcium (%)	0229	1.370	0.0200	1.349	0.0821	0.0364	121	0.26	019.41	0
019	Calcium (%)	0142	1.374	0.0434	1.349	0.0821	0.0364	121	0.30	019.31	0
019	Calcium (%)	0692	1.375	0.0500	1.349	0.0821	0.0364	121	0.32	019.42	0
019	Calcium (%)	0870	1.377	0.0130	1.349	0.0821	0.0364	121	0.34	019.43	0
019	Calcium (%)	0043	1.388	0.0250	1.349	0.0821	0.0364	121	0.47	019.00	0
019	Calcium (%)	0074	1.395	0.0100	1.349	0.0821	0.0364	121	0.56	019.41	0
019	Calcium (%)	0619	1.395	0.0100	1.349	0.0821	0.0364	121	0.56	019.41	0
019	Calcium (%)	0042	1.395	0.0300	1.349	0.0821	0.0364	121	0.56	019.42	0
019	Calcium (%)	0505	1.395	0.0100	1.349	0.0821	0.0364	121	0.56	019.43	0
019	Calcium (%)	0968	1.395	0.0040	1.349	0.0821	0.0364	121	0.56	019.43	0
019	Calcium (%)	2192	1.395	0.0700	1.349	0.0821	0.0364	121	0.56	019.43	0
019	Calcium (%)	0006	1.396	0.0360	1.349	0.0821	0.0364	121	0.57	019.43	0
019	Calcium (%)	0144	1.400	0.0800	1.349	0.0821	0.0364	121	0.62	019.41	0
019	Calcium (%)	0910	1.400	0.0800	1.349	0.0821	0.0364	121	0.62	019.41	0
019	Calcium (%)	0035	1.400	0.0000	1.349	0.0821	0.0364	121	0.62	019.42	0
019	Calcium (%)	0297	1.400	0.1000	1.349	0.0821	0.0364	121	0.62	019.43	0
019	Calcium (%)	2190	1.401	0.0026	1.349	0.0821	0.0364	121	0.63	019.08	0
019	Calcium (%)	0598	1.402	0.0037	1.349	0.0821	0.0364	121	0.64	019.41	0
019	Calcium (%)	0037	1.404	0.0220	1.349	0.0821	0.0364	121	0.67	019.42	0
019	Calcium (%)	2144	1.405	0.0100	1.349	0.0821	0.0364	121	0.68	019.00	0
019	Calcium (%)	0100	1.405	0.0100	1.349	0.0821	0.0364	121	0.68	019.99	0
019	Calcium (%)	0425	1.405	0.0700	1.349	0.0821	0.0364	121	0.68	019.43	0
019	Calcium (%)	0345	1.408	0.0190	1.349	0.0821	0.0364	121	0.72	019.43	0
019	Calcium (%)	0358	1.410	0.0000	1.349	0.0821	0.0364	121	0.75	019.41	0
019	Calcium (%)	0683	1.415	0.0100	1.349	0.0821	0.0364	121	0.81	019.00	0
019	Calcium (%)	2199	1.415	0.0500	1.349	0.0821	0.0364	121	0.81	019.08	0
019	Calcium (%)	0010	1.415	0.0100	1.349	0.0821	0.0364	121	0.81	019.43	0
019	Calcium (%)	0353	1.415	0.1700	1.349	0.0821	0.0364	121	0.81	019.43	0
019	Calcium (%)	0049	1.420	0.0400	1.349	0.0821	0.0364	121	0.87	019.41	0
019	Calcium (%)	0098	1.420	0.0000	1.349	0.0821	0.0364	121	0.87	019.43	0
019	Calcium (%)	0693	1.422	0.0670	1.349	0.0821	0.0364	121	0.89	019.42	0
019	Calcium (%)	0520	1.423	0.0420	1.349	0.0821	0.0364	121	0.90	019.41	0
019	Calcium (%)	0148	1.425	0.0100	1.349	0.0821	0.0364	121	0.93	019.41	0
019	Calcium (%)	0035	1.425	0.0100	1.349	0.0821	0.0364	121	0.93	019.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
019	Calcium (%)	0687	1.425	0.1500	1.349	0.0821	0.0364	121	0.93	019.31	0
019	Calcium (%)	0357	1.430	0.0164	1.349	0.0821	0.0364	121	0.99	019.42	0
019	Calcium (%)	0038	1.430	0.0200	1.349	0.0821	0.0364	121	0.99	019.31	0
019	Calcium (%)	0685	1.430	0.0400	1.349	0.0821	0.0364	121	0.99	019.41	0
019	Calcium (%)	0366	1.430	0.0400	1.349	0.0821	0.0364	121	0.99	019.42	0
019	Calcium (%)	0226	1.435	0.0300	1.349	0.0821	0.0364	121	1.05	019.41	0
019	Calcium (%)	0682	1.440	0.0000	1.349	0.0821	0.0364	121	1.11	019.43	0
019	Calcium (%)	2188	1.500	0.0000	1.349	0.0821	0.0364	121	1.84	019.08	0
019	Calcium (%)	0011	1.505	0.1900	1.349	0.0821	0.0364	121	1.90	019.43	0
019	Calcium (%)	0190	1.540	0.0000	1.349	0.0821	0.0364	121	2.33	019.42	0
019	Calcium (%)	0001	1.544	0.0500	1.349	0.0821	0.0364	121	2.38	019.31	0
019	Calcium (%)	0066	1.545	0.0460	1.349	0.0821	0.0364	121	2.39	019.41	0
019	Calcium (%)	0726	1.585	0.0100	1.349	0.0821	0.0364	121	2.88	019.42	0
019	Calcium (%)	0018	1.610	0.1200	1.349	0.0821	0.0364	121	3.18	019.31	0
019	Calcium (%)	0563	1.695	0.0554	1.349	0.0821	0.0364	121	4.21	019.31	0
019	Calcium (%)	0354	2.005	0.0100	1.349	0.0821	0.0364	121	7.99	019.41	0
019	Calcium (%)	0300	1.350	0.2400	1.349	0.0821	0.0364	121	0.01	019.43	1
019	Calcium (%)	2193	1.775	0.2100	1.349	0.0821	0.0364	121	5.19	019.08	1
021	Cobalt (ppm)	0619	0.9840	0.0220	1.534	0.2828	0.0405	15	-1.94	021.41	0
021	Cobalt (ppm)	0910	1.115	0.0300	1.534	0.2828	0.0405	15	-1.48	021.52	0
021	Cobalt (ppm)	0171	1.322	0.0250	1.534	0.2828	0.0405	15	-0.75	021.41	0
021	Cobalt (ppm)	0186	1.355	0.1300	1.534	0.2828	0.0405	15	-0.63	021.52	0
021	Cobalt (ppm)	0164	1.400	0.0000	1.534	0.2828	0.0405	15	-0.47	021.41	0
021	Cobalt (ppm)	0345	1.415	0.0300	1.534	0.2828	0.0405	15	-0.42	021.43	0
021	Cobalt (ppm)	0407	1.574	0.0645	1.534	0.2828	0.0405	15	0.14	021.53	0
021	Cobalt (ppm)	0939	1.575	0.0900	1.534	0.2828	0.0405	15	0.15	021.53	0
021	Cobalt (ppm)	0407	1.595	0.0499	1.534	0.2828	0.0405	15	0.22	021.43	0
021	Cobalt (ppm)	0169	1.615	0.0100	1.534	0.2828	0.0405	15	0.29	021.43	0
021	Cobalt (ppm)	0407	1.637	0.0055	1.534	0.2828	0.0405	15	0.37	021.41	0
021	Cobalt (ppm)	0510	1.705	0.1300	1.534	0.2828	0.0405	15	0.61	021.43	0
021	Cobalt (ppm)	0553	1.730	0.0200	1.534	0.2828	0.0405	15	0.69	021.53	0
021	Cobalt (ppm)	0045	1.900	0.0000	1.534	0.2828	0.0405	15	1.30	021.42	0
021	Cobalt (ppm)	0682	2.500	0.0000	1.534	0.2828	0.0405	15	3.42	021.43	0
021	Cobalt (ppm)	0353	0.6183	0.2942	1.534	0.2828	0.0405	15	-3.24	021.43	1
022	Copper (ppm)	0536	15.55	0.7000	35.22	2.910	1.196	78	-6.76	022.31	0
022	Copper (ppm)	0598	25.56	1.288	35.22	2.910	1.196	78	-3.32	022.41	0
022	Copper (ppm)	0049	27.50	1.680	35.22	2.910	1.196	78	-2.65	022.41	0
022	Copper (ppm)	0619	27.80	1.800	35.22	2.910	1.196	78	-2.55	022.41	0
022	Copper (ppm)	0629	28.50	1.000	35.22	2.910	1.196	78	-2.31	022.31	0
022	Copper (ppm)	0229	29.65	1.270	35.22	2.910	1.196	78	-1.92	022.41	0
022	Copper (ppm)	0563	30.43	0.6900	35.22	2.910	1.196	78	-1.65	022.31	0
022	Copper (ppm)	0590	30.95	0.9000	35.22	2.910	1.196	78	-1.47	022.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
022	Copper (ppm)	0098	31.29	2.570	35.22	2.910	1.196	78	-1.35	022.41	0
022	Copper (ppm)	0407	31.59	1.101	35.22	2.910	1.196	78	-1.25	022.53	0
022	Copper (ppm)	0910	32.00	2.000	35.22	2.910	1.196	78	-1.11	022.41	0
022	Copper (ppm)	0263	32.32	0.3370	35.22	2.910	1.196	78	-1.00	022.42	0
022	Copper (ppm)	0358	32.41	0.1200	35.22	2.910	1.196	78	-0.97	022.41	0
022	Copper (ppm)	0407	32.69	1.613	35.22	2.910	1.196	78	-0.87	022.41	0
022	Copper (ppm)	0553	32.75	1.500	35.22	2.910	1.196	78	-0.85	022.53	0
022	Copper (ppm)	0511	33.00	2.000	35.22	2.910	1.196	78	-0.76	022.43	0
022	Copper (ppm)	0045	33.30	0.4000	35.22	2.910	1.196	78	-0.66	022.42	0
022	Copper (ppm)	0169	33.30	0.2000	35.22	2.910	1.196	78	-0.66	022.43	0
022	Copper (ppm)	0004	33.38	0.8100	35.22	2.910	1.196	78	-0.63	022.41	0
022	Copper (ppm)	0674	33.50	0.0000	35.22	2.910	1.196	78	-0.59	022.41	0
022	Copper (ppm)	0083	33.50	1.000	35.22	2.910	1.196	78	-0.59	022.43	0
022	Copper (ppm)	2129	33.72	0.5200	35.22	2.910	1.196	78	-0.52	022.42	0
022	Copper (ppm)	0425	33.77	0.5300	35.22	2.910	1.196	78	-0.50	022.43	0
022	Copper (ppm)	0653	33.77	0.1800	35.22	2.910	1.196	78	-0.50	022.41	0
022	Copper (ppm)	0171	33.97	0.2800	35.22	2.910	1.196	78	-0.43	022.41	0
022	Copper (ppm)	0297	34.00	0.0000	35.22	2.910	1.196	78	-0.42	022.43	0
022	Copper (ppm)	0100	34.00	0.0000	35.22	2.910	1.196	78	-0.42	022.99	0
022	Copper (ppm)	0066	34.24	2.460	35.22	2.910	1.196	78	-0.34	022.44	0
022	Copper (ppm)	0047	34.55	0.0300	35.22	2.910	1.196	78	-0.23	022.52	0
022	Copper (ppm)	0682	34.70	0.0000	35.22	2.910	1.196	78	-0.18	022.43	0
022	Copper (ppm)	2246	34.74	0.4900	35.22	2.910	1.196	78	-0.17	022.31	0
022	Copper (ppm)	0939	34.74	1.760	35.22	2.910	1.196	78	-0.17	022.53	0
022	Copper (ppm)	0407	34.82	4.874	35.22	2.910	1.196	78	-0.14	022.43	0
022	Copper (ppm)	2319	34.85	0.9000	35.22	2.910	1.196	78	-0.13	022.42	0
022	Copper (ppm)	0226	34.88	0.6700	35.22	2.910	1.196	78	-0.12	022.41	0
022	Copper (ppm)	0035	35.00	2.000	35.22	2.910	1.196	78	-0.08	022.43	0
022	Copper (ppm)	0164	35.00	0.0000	35.22	2.910	1.196	78	-0.08	022.44	0
022	Copper (ppm)	0918	35.13	0.5300	35.22	2.910	1.196	78	-0.03	022.43	0
022	Copper (ppm)	0098	35.16	1.310	35.22	2.910	1.196	78	-0.02	022.43	0
022	Copper (ppm)	0148	35.18	4.002	35.22	2.910	1.196	78	-0.01	022.41	0
022	Copper (ppm)	0017	35.18	3.070	35.22	2.910	1.196	78	-0.01	022.43	0
022	Copper (ppm)	2196	35.20	0.0000	35.22	2.910	1.196	78	-0.01	022.31	0
022	Copper (ppm)	0366	35.50	3.000	35.22	2.910	1.196	78	0.10	022.42	0
022	Copper (ppm)	0675	35.50	0.0800	35.22	2.910	1.196	78	0.10	022.43	0
022	Copper (ppm)	0692	35.70	2.800	35.22	2.910	1.196	78	0.16	022.42	0
022	Copper (ppm)	0186	35.70	0.8000	35.22	2.910	1.196	78	0.16	022.52	0
022	Copper (ppm)	0010	35.85	2.100	35.22	2.910	1.196	78	0.22	022.33	0
022	Copper (ppm)	0278	35.95	2.500	35.22	2.910	1.196	78	0.25	022.42	0
022	Copper (ppm)	0520	36.05	0.6960	35.22	2.910	1.196	78	0.28	022.41	0
022	Copper (ppm)	0265	36.10	0.8000	35.22	2.910	1.196	78	0.30	022.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
022	Copper (ppm)	0848	36.15	2.280	35.22	2.910	1.196	78	0.32	022.41	0
022	Copper (ppm)	0689	36.15	1.100	35.22	2.910	1.196	78	0.32	022.31	0
022	Copper (ppm)	0726	36.33	0.6600	35.22	2.910	1.196	78	0.38	022.42	0
022	Copper (ppm)	0510	36.50	1.000	35.22	2.910	1.196	78	0.44	022.43	0
022	Copper (ppm)	0345	36.56	0.2800	35.22	2.910	1.196	78	0.46	022.43	0
022	Copper (ppm)	0572	36.65	1.300	35.22	2.910	1.196	78	0.49	022.53	0
022	Copper (ppm)	0870	36.73	0.7300	35.22	2.910	1.196	78	0.52	022.42	0
022	Copper (ppm)	0043	36.82	1.611	35.22	2.910	1.196	78	0.55	022.43	0
022	Copper (ppm)	0038	37.05	3.900	35.22	2.910	1.196	78	0.63	022.32	0
022	Copper (ppm)	0027	37.05	0.0820	35.22	2.910	1.196	78	0.63	022.43	0
022	Copper (ppm)	0693	37.33	2.438	35.22	2.910	1.196	78	0.73	022.42	0
022	Copper (ppm)	0186	37.35	0.1000	35.22	2.910	1.196	78	0.73	022.42	0
022	Copper (ppm)	0009	37.50	5.000	35.22	2.910	1.196	78	0.78	022.42	0
022	Copper (ppm)	0037	37.80	2.000	35.22	2.910	1.196	78	0.89	022.42	0
022	Copper (ppm)	0035	38.00	0.0000	35.22	2.910	1.196	78	0.96	022.42	0
022	Copper (ppm)	0968	38.02	0.0940	35.22	2.910	1.196	78	0.96	022.43	0
022	Copper (ppm)	0019	38.19	0.0800	35.22	2.910	1.196	78	1.02	022.41	0
022	Copper (ppm)	0357	38.37	0.0240	35.22	2.910	1.196	78	1.08	022.42	0
022	Copper (ppm)	0190	38.43	0.1000	35.22	2.910	1.196	78	1.10	022.42	0
022	Copper (ppm)	0870	38.93	0.7500	35.22	2.910	1.196	78	1.27	022.43	0
022	Copper (ppm)	0685	39.41	0.1600	35.22	2.910	1.196	78	1.44	022.42	0
022	Copper (ppm)	0353	39.50	1.000	35.22	2.910	1.196	78	1.47	022.43	0
022	Copper (ppm)	0505	39.63	2.570	35.22	2.910	1.196	78	1.51	022.43	0
022	Copper (ppm)	0042	40.50	2.800	35.22	2.910	1.196	78	1.81	022.42	0
022	Copper (ppm)	0529	41.75	0.5000	35.22	2.910	1.196	78	2.24	022.31	0
022	Copper (ppm)	0651	43.02	0.1016	35.22	2.910	1.196	78	2.68	022.31	0
022	Copper (ppm)	2089	43.32	3.230	35.22	2.910	1.196	78	2.78	022.43	0
022	Copper (ppm)	2188	44.10	0.0000	35.22	2.910	1.196	78	3.05	022.31	0
022	Copper (ppm)	0354	73.45	0.4700	35.22	2.910	1.196	78	13.14	022.41	2
025	Iron (ppm)	0037	184.0	29.40	345.2	39.99	8.684	63	-4.03	025.42	0
025	Iron (ppm)	0265	230.5	7.000	345.2	39.99	8.684	63	-2.87	025.42	0
025	Iron (ppm)	2319	242.5	1.000	345.2	39.99	8.684	63	-2.57	025.42	0
025	Iron (ppm)	0536	256.5	1.000	345.2	39.99	8.684	63	-2.22	025.31	0
025	Iron (ppm)	0692	265.1	5.200	345.2	39.99	8.684	63	-2.00	025.42	0
025	Iron (ppm)	0190	266.1	0.4000	345.2	39.99	8.684	63	-1.98	025.42	0
025	Iron (ppm)	0278	273.0	2.000	345.2	39.99	8.684	63	-1.81	025.42	0
025	Iron (ppm)	0726	279.1	4.100	345.2	39.99	8.684	63	-1.65	025.42	0
025	Iron (ppm)	2129	280.6	2.900	345.2	39.99	8.684	63	-1.62	025.42	0
025	Iron (ppm)	0425	285.3	23.30	345.2	39.99	8.684	63	-1.50	025.43	0
025	Iron (ppm)	0045	299.5	7.000	345.2	39.99	8.684	63	-1.14	025.42	0
025	Iron (ppm)	0357	301.1	0.5420	345.2	39.99	8.684	63	-1.10	025.42	0
025	Iron (ppm)	0870	302.0	8.100	345.2	39.99	8.684	63	-1.08	025.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
025	Iron (ppm)	0345	312.0	6.000	345.2	39.99	8.684	63	-0.83	025.43	0
025	Iron (ppm)	0083	313.5	3.000	345.2	39.99	8.684	63	-0.79	025.43	0
025	Iron (ppm)	0968	320.1	2.440	345.2	39.99	8.684	63	-0.63	025.43	0
025	Iron (ppm)	0653	321.4	2.700	345.2	39.99	8.684	63	-0.60	025.41	0
025	Iron (ppm)	0366	327.5	19.00	345.2	39.99	8.684	63	-0.44	025.42	0
025	Iron (ppm)	0263	332.6	5.960	345.2	39.99	8.684	63	-0.32	025.42	0
025	Iron (ppm)	0123	336.5	11.00	345.2	39.99	8.684	63	-0.22	025.41	0
025	Iron (ppm)	0358	341.3	3.730	345.2	39.99	8.684	63	-0.10	025.41	0
025	Iron (ppm)	0098	342.8	25.50	345.2	39.99	8.684	63	-0.06	025.41	0
025	Iron (ppm)	0598	343.0	3.955	345.2	39.99	8.684	63	-0.06	025.41	0
025	Iron (ppm)	0297	344.5	15.00	345.2	39.99	8.684	63	-0.02	025.43	0
025	Iron (ppm)	0511	345.0	10.00	345.2	39.99	8.684	63	-0.01	025.43	0
025	Iron (ppm)	0407	346.1	10.64	345.2	39.99	8.684	63	0.02	025.43	0
025	Iron (ppm)	0169	347.0	2.000	345.2	39.99	8.684	63	0.04	025.43	0
025	Iron (ppm)	0407	348.6	5.230	345.2	39.99	8.684	63	0.08	025.53	0
025	Iron (ppm)	0226	348.7	1.410	345.2	39.99	8.684	63	0.09	025.41	0
025	Iron (ppm)	2089	348.9	20.47	345.2	39.99	8.684	63	0.09	025.43	0
025	Iron (ppm)	0553	351.0	16.00	345.2	39.99	8.684	63	0.14	025.53	0
025	Iron (ppm)	0505	351.9	21.80	345.2	39.99	8.684	63	0.17	025.43	0
025	Iron (ppm)	0407	354.2	6.581	345.2	39.99	8.684	63	0.22	025.41	0
025	Iron (ppm)	0520	354.5	11.43	345.2	39.99	8.684	63	0.23	025.41	0
025	Iron (ppm)	0049	355.9	8.090	345.2	39.99	8.684	63	0.27	025.41	0
025	Iron (ppm)	0939	358.0	10.00	345.2	39.99	8.684	63	0.32	025.53	0
025	Iron (ppm)	0870	359.1	7.800	345.2	39.99	8.684	63	0.35	025.43	0
025	Iron (ppm)	0910	360.5	13.00	345.2	39.99	8.684	63	0.38	025.41	0
025	Iron (ppm)	0074	361.5	7.000	345.2	39.99	8.684	63	0.41	025.41	0
025	Iron (ppm)	0693	362.0	0.8800	345.2	39.99	8.684	63	0.42	025.42	0
025	Iron (ppm)	2196	362.4	0.0000	345.2	39.99	8.684	63	0.43	025.31	0
025	Iron (ppm)	0004	364.0	9.250	345.2	39.99	8.684	63	0.47	025.41	0
025	Iron (ppm)	0682	364.8	0.0000	345.2	39.99	8.684	63	0.49	025.43	0
025	Iron (ppm)	0689	364.8	13.40	345.2	39.99	8.684	63	0.49	025.31	0
025	Iron (ppm)	0229	365.0	27.30	345.2	39.99	8.684	63	0.49	025.41	0
025	Iron (ppm)	0529	365.8	6.400	345.2	39.99	8.684	63	0.51	025.31	0
025	Iron (ppm)	0038	368.5	17.00	345.2	39.99	8.684	63	0.58	025.31	0
025	Iron (ppm)	0675	370.3	0.5500	345.2	39.99	8.684	63	0.63	025.43	0
025	Iron (ppm)	0876	370.5	5.000	345.2	39.99	8.684	63	0.63	025.43	0
025	Iron (ppm)	0510	371.0	6.000	345.2	39.99	8.684	63	0.64	025.43	0
025	Iron (ppm)	0100	371.5	13.00	345.2	39.99	8.684	63	0.66	025.99	0
025	Iron (ppm)	0164	373.5	13.00	345.2	39.99	8.684	63	0.71	025.41	0
025	Iron (ppm)	0171	376.3	3.300	345.2	39.99	8.684	63	0.78	025.41	0
025	Iron (ppm)	0918	376.4	10.24	345.2	39.99	8.684	63	0.78	025.43	0
025	Iron (ppm)	0651	382.5	1.219	345.2	39.99	8.684	63	0.93	025.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
025	Iron (ppm)	0035	384.5	3.000	345.2	39.99	8.684	63	0.98	025.43	0
025	Iron (ppm)	0035	386.0	24.00	345.2	39.99	8.684	63	1.02	025.42	0
025	Iron (ppm)	2022	390.0	6.000	345.2	39.99	8.684	63	1.12	025.31	0
025	Iron (ppm)	0098	391.2	8.200	345.2	39.99	8.684	63	1.15	025.43	0
025	Iron (ppm)	0148	391.7	0.8700	345.2	39.99	8.684	63	1.16	025.41	0
025	Iron (ppm)	2188	407.3	0.0000	345.2	39.99	8.684	63	1.55	025.31	0
025	Iron (ppm)	0353	427.0	36.00	345.2	39.99	8.684	63	2.04	025.43	0
025	Iron (ppm)	0354	523.7	0.8100	345.2	39.99	8.684	63	4.46	025.41	0
025	Iron (ppm)	0042	300.0	72.00	345.2	39.99	8.684	63	-1.13	025.42	1
025	Iron (ppm)	2246	378.7	56.23	345.2	39.99	8.684	63	0.84	025.31	1
025	Iron (ppm)	0563	864.0	17.81	345.2	39.99	8.684	63	12.97	025.31	2
027	Magnesium (%)	2129	0.1703	0.0001	0.2030	0.0096	0.0052	66	-3.39	027.42	0
027	Magnesium (%)	0590	0.1800	0.0000	0.2030	0.0096	0.0052	66	-2.38	027.99	0
027	Magnesium (%)	0407	0.1838	0.0019	0.2030	0.0096	0.0052	66	-1.99	027.43	0
027	Magnesium (%)	0083	0.1845	0.0006	0.2030	0.0096	0.0052	66	-1.91	027.43	0
027	Magnesium (%)	0889	0.1850	0.0100	0.2030	0.0096	0.0052	66	-1.86	027.99	0
027	Magnesium (%)	0407	0.1858	0.0058	0.2030	0.0096	0.0052	66	-1.78	027.41	0
027	Magnesium (%)	0407	0.1864	0.0047	0.2030	0.0096	0.0052	66	-1.72	027.53	0
027	Magnesium (%)	0045	0.1895	0.0050	0.2030	0.0096	0.0052	66	-1.39	027.42	0
027	Magnesium (%)	0049	0.1900	0.0000	0.2030	0.0096	0.0052	66	-1.34	027.41	0
027	Magnesium (%)	0939	0.1900	0.0000	0.2030	0.0096	0.0052	66	-1.34	027.53	0
027	Magnesium (%)	0004	0.1950	0.0100	0.2030	0.0096	0.0052	66	-0.82	027.41	0
027	Magnesium (%)	0098	0.1950	0.0100	0.2030	0.0096	0.0052	66	-0.82	027.41	0
027	Magnesium (%)	0229	0.1950	0.0100	0.2030	0.0096	0.0052	66	-0.82	027.41	0
027	Magnesium (%)	0685	0.1950	0.0100	0.2030	0.0096	0.0052	66	-0.82	027.42	0
027	Magnesium (%)	0870	0.1956	0.0018	0.2030	0.0096	0.0052	66	-0.76	027.42	0
027	Magnesium (%)	0870	0.1963	0.0033	0.2030	0.0096	0.0052	66	-0.69	027.43	0
027	Magnesium (%)	0876	0.1970	0.0120	0.2030	0.0096	0.0052	66	-0.62	027.43	0
027	Magnesium (%)	0038	0.1975	0.0010	0.2030	0.0096	0.0052	66	-0.57	027.31	0
027	Magnesium (%)	0186	0.1987	0.0019	0.2030	0.0096	0.0052	66	-0.45	027.52	0
027	Magnesium (%)	0006	0.1988	0.0024	0.2030	0.0096	0.0052	66	-0.43	027.43	0
027	Magnesium (%)	0154	0.1996	0.0037	0.2030	0.0096	0.0052	66	-0.35	027.52	0
027	Magnesium (%)	0968	0.1998	0.0007	0.2030	0.0096	0.0052	66	-0.33	027.43	0
027	Magnesium (%)	0358	0.2000	0.0000	0.2030	0.0096	0.0052	66	-0.31	027.41	0
027	Magnesium (%)	0692	0.2000	0.0000	0.2030	0.0096	0.0052	66	-0.31	027.42	0
027	Magnesium (%)	2319	0.2000	0.0000	0.2030	0.0096	0.0052	66	-0.31	027.42	0
027	Magnesium (%)	0510	0.2000	0.0000	0.2030	0.0096	0.0052	66	-0.31	027.43	0
027	Magnesium (%)	0675	0.2000	0.0000	0.2030	0.0096	0.0052	66	-0.31	027.43	0
027	Magnesium (%)	2089	0.2000	0.0000	0.2030	0.0096	0.0052	66	-0.31	027.43	0
027	Magnesium (%)	0142	0.2018	0.0137	0.2030	0.0096	0.0052	66	-0.12	027.31	0
027	Magnesium (%)	0263	0.2030	0.0000	0.2030	0.0096	0.0052	66	0.00	027.42	0
027	Magnesium (%)	0164	0.2040	0.0080	0.2030	0.0096	0.0052	66	0.11	027.44	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
027	Magnesium (%)	0357	0.2042	0.0007	0.2030	0.0096	0.0052	66	0.12	027.42	0
027	Magnesium (%)	0042	0.2050	0.0020	0.2030	0.0096	0.0052	66	0.21	027.42	0
027	Magnesium (%)	0918	0.2050	0.0000	0.2030	0.0096	0.0052	66	0.21	027.43	0
027	Magnesium (%)	0265	0.2050	0.0100	0.2030	0.0096	0.0052	66	0.21	027.42	0
027	Magnesium (%)	0366	0.2050	0.0100	0.2030	0.0096	0.0052	66	0.21	027.42	0
027	Magnesium (%)	0726	0.2050	0.0100	0.2030	0.0096	0.0052	66	0.21	027.42	0
027	Magnesium (%)	0345	0.2050	0.0080	0.2030	0.0096	0.0052	66	0.21	027.43	0
027	Magnesium (%)	0425	0.2050	0.0100	0.2030	0.0096	0.0052	66	0.21	027.43	0
027	Magnesium (%)	0100	0.2050	0.0100	0.2030	0.0096	0.0052	66	0.21	027.99	0
027	Magnesium (%)	0598	0.2052	0.0017	0.2030	0.0096	0.0052	66	0.23	027.41	0
027	Magnesium (%)	0171	0.2060	0.0020	0.2030	0.0096	0.0052	66	0.32	027.41	0
027	Magnesium (%)	0563	0.2069	0.0031	0.2030	0.0096	0.0052	66	0.40	027.31	0
027	Magnesium (%)	0572	0.2070	0.0020	0.2030	0.0096	0.0052	66	0.42	027.53	0
027	Magnesium (%)	0037	0.2075	0.0030	0.2030	0.0096	0.0052	66	0.47	027.42	0
027	Magnesium (%)	0553	0.2075	0.0030	0.2030	0.0096	0.0052	66	0.47	027.53	0
027	Magnesium (%)	2022	0.2080	0.0020	0.2030	0.0096	0.0052	66	0.52	027.31	0
027	Magnesium (%)	0074	0.2085	0.0050	0.2030	0.0096	0.0052	66	0.57	027.41	0
027	Magnesium (%)	0653	0.2095	0.0010	0.2030	0.0096	0.0052	66	0.68	027.41	0
027	Magnesium (%)	0226	0.2100	0.0000	0.2030	0.0096	0.0052	66	0.73	027.41	0
027	Magnesium (%)	0035	0.2100	0.0000	0.2030	0.0096	0.0052	66	0.73	027.42	0
027	Magnesium (%)	0278	0.2100	0.0000	0.2030	0.0096	0.0052	66	0.73	027.42	0
027	Magnesium (%)	0035	0.2100	0.0000	0.2030	0.0096	0.0052	66	0.73	027.43	0
027	Magnesium (%)	0098	0.2100	0.0000	0.2030	0.0096	0.0052	66	0.73	027.43	0
027	Magnesium (%)	0682	0.2100	0.0000	0.2030	0.0096	0.0052	66	0.73	027.43	0
027	Magnesium (%)	0169	0.2100	0.0400	0.2030	0.0096	0.0052	66	0.73	027.32	0
027	Magnesium (%)	0297	0.2100	0.0200	0.2030	0.0096	0.0052	66	0.73	027.43	0
027	Magnesium (%)	0148	0.2115	0.0110	0.2030	0.0096	0.0052	66	0.89	027.41	0
027	Magnesium (%)	0520	0.2130	0.0020	0.2030	0.0096	0.0052	66	1.04	027.41	0
027	Magnesium (%)	0910	0.2150	0.0100	0.2030	0.0096	0.0052	66	1.25	027.41	0
027	Magnesium (%)	2196	0.2200	0.0000	0.2030	0.0096	0.0052	66	1.77	027.31	0
027	Magnesium (%)	0186	0.2243	0.0110	0.2030	0.0096	0.0052	66	2.21	027.42	0
027	Magnesium (%)	0693	0.2325	0.0090	0.2030	0.0096	0.0052	66	3.06	027.42	0
027	Magnesium (%)	0353	0.2350	0.0100	0.2030	0.0096	0.0052	66	3.32	027.43	0
027	Magnesium (%)	2302	0.2400	0.0200	0.2030	0.0096	0.0052	66	3.84	027.99	0
027	Magnesium (%)	0354	0.3350	0.0100	0.2030	0.0096	0.0052	66	13.68	027.41	0
027	Magnesium (%)	0536	25.20	0.2000	0.2030	0.0096	0.0052	66	2590.43	027.31	2
028	Manganese (ppm)	0689	86.70	9.800	118.3	7.608	3.711	78	-4.15	028.31	0
028	Manganese (ppm)	0619	90.50	3.200	118.3	7.608	3.711	78	-3.65	028.41	0
028	Manganese (ppm)	0226	95.62	1.720	118.3	7.608	3.711	78	-2.98	028.41	0
028	Manganese (ppm)	2129	100.8	0.9000	118.3	7.608	3.711	78	-2.30	028.42	0
028	Manganese (ppm)	0870	105.1	5.900	118.3	7.608	3.711	78	-1.74	028.43	0
028	Manganese (ppm)	0870	107.4	3.100	118.3	7.608	3.711	78	-1.43	028.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
028	Manganese (ppm)	0047	107.8	12.45	118.3	7.608	3.711	78	-1.37	028.52	0
028	Manganese (ppm)	0590	108.5	3.000	118.3	7.608	3.711	78	-1.28	028.99	0
028	Manganese (ppm)	0553	109.0	12.00	118.3	7.608	3.711	78	-1.22	028.53	0
028	Manganese (ppm)	0685	109.0	1.400	118.3	7.608	3.711	78	-1.21	028.42	0
028	Manganese (ppm)	0653	110.5	0.6000	118.3	7.608	3.711	78	-1.02	028.41	0
028	Manganese (ppm)	0083	110.5	1.000	118.3	7.608	3.711	78	-1.02	028.43	0
028	Manganese (ppm)	0049	111.5	1.140	118.3	7.608	3.711	78	-0.89	028.41	0
028	Manganese (ppm)	0263	111.9	0.6570	118.3	7.608	3.711	78	-0.83	028.42	0
028	Manganese (ppm)	0169	112.0	2.000	118.3	7.608	3.711	78	-0.82	028.43	0
028	Manganese (ppm)	2196	112.4	0.0000	118.3	7.608	3.711	78	-0.77	028.31	0
028	Manganese (ppm)	0505	113.2	2.000	118.3	7.608	3.711	78	-0.66	028.43	0
028	Manganese (ppm)	0074	113.5	3.000	118.3	7.608	3.711	78	-0.62	028.41	0
028	Manganese (ppm)	0045	113.5	13.00	118.3	7.608	3.711	78	-0.62	028.42	0
028	Manganese (ppm)	0511	113.5	1.000	118.3	7.608	3.711	78	-0.62	028.43	0
028	Manganese (ppm)	0066	113.7	5.200	118.3	7.608	3.711	78	-0.60	028.44	0
028	Manganese (ppm)	0098	113.8	0.4000	118.3	7.608	3.711	78	-0.59	028.41	0
028	Manganese (ppm)	0358	114.1	1.130	118.3	7.608	3.711	78	-0.55	028.41	0
028	Manganese (ppm)	0407	114.4	2.766	118.3	7.608	3.711	78	-0.50	028.53	0
028	Manganese (ppm)	0164	114.5	1.000	118.3	7.608	3.711	78	-0.49	028.44	0
028	Manganese (ppm)	0939	115.0	2.000	118.3	7.608	3.711	78	-0.43	028.53	0
028	Manganese (ppm)	0019	115.2	1.300	118.3	7.608	3.711	78	-0.41	028.41	0
028	Manganese (ppm)	0563	115.2	0.3535	118.3	7.608	3.711	78	-0.41	028.31	0
028	Manganese (ppm)	0598	115.5	2.013	118.3	7.608	3.711	78	-0.37	028.41	0
028	Manganese (ppm)	0693	115.5	6.361	118.3	7.608	3.711	78	-0.36	028.42	0
028	Manganese (ppm)	0009	115.5	7.000	118.3	7.608	3.711	78	-0.36	028.42	0
028	Manganese (ppm)	0918	115.7	4.340	118.3	7.608	3.711	78	-0.34	028.43	0
028	Manganese (ppm)	0407	116.5	5.285	118.3	7.608	3.711	78	-0.23	028.43	0
028	Manganese (ppm)	0229	116.7	1.900	118.3	7.608	3.711	78	-0.21	028.41	0
028	Manganese (ppm)	0407	116.8	7.315	118.3	7.608	3.711	78	-0.19	028.41	0
028	Manganese (ppm)	0910	117.5	5.000	118.3	7.608	3.711	78	-0.10	028.41	0
028	Manganese (ppm)	0278	117.5	13.00	118.3	7.608	3.711	78	-0.10	028.42	0
028	Manganese (ppm)	0572	117.5	5.000	118.3	7.608	3.711	78	-0.10	028.53	0
028	Manganese (ppm)	0651	117.5	0.1089	118.3	7.608	3.711	78	-0.09	028.31	0
028	Manganese (ppm)	2319	118.0	2.000	118.3	7.608	3.711	78	-0.03	028.42	0
028	Manganese (ppm)	0425	118.9	1.000	118.3	7.608	3.711	78	0.08	028.43	0
028	Manganese (ppm)	0510	119.0	0.0000	118.3	7.608	3.711	78	0.10	028.43	0
028	Manganese (ppm)	2188	119.2	0.0000	118.3	7.608	3.711	78	0.13	028.31	0
028	Manganese (ppm)	0001	119.4	0.7000	118.3	7.608	3.711	78	0.14	028.31	0
028	Manganese (ppm)	0043	119.4	14.43	118.3	7.608	3.711	78	0.15	028.43	0
028	Manganese (ppm)	0038	119.5	9.000	118.3	7.608	3.711	78	0.16	028.32	0
028	Manganese (ppm)	0098	120.0	2.500	118.3	7.608	3.711	78	0.22	028.43	0
028	Manganese (ppm)	0529	120.4	0.1000	118.3	7.608	3.711	78	0.28	028.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
028	Manganese (ppm)	2022	120.5	3.000	118.3	7.608	3.711	78	0.30	028.31	0
028	Manganese (ppm)	0876	120.5	5.000	118.3	7.608	3.711	78	0.30	028.43	0
028	Manganese (ppm)	0520	120.7	4.003	118.3	7.608	3.711	78	0.33	028.41	0
028	Manganese (ppm)	0186	120.9	6.300	118.3	7.608	3.711	78	0.34	028.52	0
028	Manganese (ppm)	0171	121.1	2.100	118.3	7.608	3.711	78	0.37	028.41	0
028	Manganese (ppm)	0675	121.3	9.090	118.3	7.608	3.711	78	0.40	028.43	0
028	Manganese (ppm)	0004	121.8	0.2100	118.3	7.608	3.711	78	0.46	028.41	0
028	Manganese (ppm)	0297	122.0	6.000	118.3	7.608	3.711	78	0.49	028.43	0
028	Manganese (ppm)	0186	122.1	6.900	118.3	7.608	3.711	78	0.50	028.42	0
028	Manganese (ppm)	0674	122.4	0.0000	118.3	7.608	3.711	78	0.55	028.41	0
028	Manganese (ppm)	0042	122.5	7.000	118.3	7.608	3.711	78	0.56	028.42	0
028	Manganese (ppm)	0366	122.5	13.00	118.3	7.608	3.711	78	0.56	028.42	0
028	Manganese (ppm)	2089	122.9	0.6600	118.3	7.608	3.711	78	0.61	028.43	0
028	Manganese (ppm)	0265	124.5	1.000	118.3	7.608	3.711	78	0.82	028.42	0
028	Manganese (ppm)	0035	125.0	2.000	118.3	7.608	3.711	78	0.89	028.42	0
028	Manganese (ppm)	0345	125.0	2.000	118.3	7.608	3.711	78	0.89	028.43	0
028	Manganese (ppm)	0692	125.2	11.40	118.3	7.608	3.711	78	0.91	028.42	0
028	Manganese (ppm)	0017	125.6	4.169	118.3	7.608	3.711	78	0.96	028.43	0
028	Manganese (ppm)	0682	126.1	0.0000	118.3	7.608	3.711	78	1.02	028.43	0
028	Manganese (ppm)	0357	127.5	0.7350	118.3	7.608	3.711	78	1.21	028.42	0
028	Manganese (ppm)	0100	128.0	2.000	118.3	7.608	3.711	78	1.28	028.99	0
028	Manganese (ppm)	0148	129.0	4.790	118.3	7.608	3.711	78	1.41	028.41	0
028	Manganese (ppm)	0190	129.4	0.3000	118.3	7.608	3.711	78	1.46	028.42	0
028	Manganese (ppm)	0037	130.1	2.900	118.3	7.608	3.711	78	1.55	028.42	0
028	Manganese (ppm)	0968	135.4	3.270	118.3	7.608	3.711	78	2.26	028.43	0
028	Manganese (ppm)	0035	137.0	6.000	118.3	7.608	3.711	78	2.46	028.43	0
028	Manganese (ppm)	0726	138.8	1.100	118.3	7.608	3.711	78	2.69	028.42	0
028	Manganese (ppm)	0353	142.0	2.000	118.3	7.608	3.711	78	3.12	028.43	0
028	Manganese (ppm)	2246	153.7	2.270	118.3	7.608	3.711	78	4.66	028.34	0
028	Manganese (ppm)	0354	165.3	1.190	118.3	7.608	3.711	78	6.19	028.41	0
028	Manganese (ppm)	0027	137.5	19.45	118.3	7.608	3.711	78	2.53	028.43	1
028	Manganese (ppm)	0536	25.30	0.6000	118.3	7.608	3.711	78	-12.22	028.31	2
030	Nitrate (%)	0357	0.0075	0.0007				1		030.01	0
031	Phosphorus (%)	0889	0.5250	0.0100	0.7287	0.0265	0.0184	117	-7.68	031.99	0
031	Phosphorus (%)	2337	0.5450	0.0100	0.7287	0.0265	0.0184	117	-6.92	031.01	0
031	Phosphorus (%)	0626	0.6000	0.0340	0.7287	0.0265	0.0184	117	-4.85	031.01	0
031	Phosphorus (%)	0629	0.6450	0.0100	0.7287	0.0265	0.0184	117	-3.15	031.01	0
031	Phosphorus (%)	2129	0.6547	0.0025	0.7287	0.0265	0.0184	117	-2.79	031.42	0
031	Phosphorus (%)	0848	0.6550	0.0100	0.7287	0.0265	0.0184	117	-2.78	031.41	0
031	Phosphorus (%)	0154	0.6576	0.0313	0.7287	0.0265	0.0184	117	-2.68	031.52	0
031	Phosphorus (%)	0689	0.6600	0.0200	0.7287	0.0265	0.0184	117	-2.59	031.01	0
031	Phosphorus (%)	0674	0.6700	0.0000	0.7287	0.0265	0.0184	117	-2.21	031.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	0300	0.6790	0.0800	0.7287	0.0265	0.0184	117	-1.87	031.43	0
031	Phosphorus (%)	0407	0.6805	0.0074	0.7287	0.0265	0.0184	117	-1.82	031.53	0
031	Phosphorus (%)	0083	0.6925	0.0010	0.7287	0.0265	0.0184	117	-1.36	031.43	0
031	Phosphorus (%)	2109	0.6930	0.0260	0.7287	0.0265	0.0184	117	-1.34	031.01	0
031	Phosphorus (%)	0019	0.6950	0.0500	0.7287	0.0265	0.0184	117	-1.27	031.41	0
031	Phosphorus (%)	2188	0.7000	0.0000	0.7287	0.0265	0.0184	117	-1.08	031.01	0
031	Phosphorus (%)	0004	0.7000	0.0000	0.7287	0.0265	0.0184	117	-1.08	031.41	0
031	Phosphorus (%)	0685	0.7000	0.0200	0.7287	0.0265	0.0184	117	-1.08	031.41	0
031	Phosphorus (%)	0572	0.7025	0.0110	0.7287	0.0265	0.0184	117	-0.99	031.53	0
031	Phosphorus (%)	0001	0.7030	0.0020	0.7287	0.0265	0.0184	117	-0.97	031.43	0
031	Phosphorus (%)	0045	0.7040	0.0020	0.7287	0.0265	0.0184	117	-0.93	031.42	0
031	Phosphorus (%)	0016	0.7050	0.0140	0.7287	0.0265	0.0184	117	-0.89	031.01	0
031	Phosphorus (%)	0939	0.7050	0.1100	0.7287	0.0265	0.0184	117	-0.89	031.01	0
031	Phosphorus (%)	2200	0.7050	0.0500	0.7287	0.0265	0.0184	117	-0.89	031.01	0
031	Phosphorus (%)	0651	0.7056	0.0056	0.7287	0.0265	0.0184	117	-0.87	031.01	0
031	Phosphorus (%)	0038	0.7070	0.0060	0.7287	0.0265	0.0184	117	-0.82	031.01	0
031	Phosphorus (%)	0407	0.7072	0.0220	0.7287	0.0265	0.0184	117	-0.81	031.41	0
031	Phosphorus (%)	0529	0.7100	0.0200	0.7287	0.0265	0.0184	117	-0.70	031.01	0
031	Phosphorus (%)	0590	0.7100	0.0000	0.7287	0.0265	0.0184	117	-0.70	031.01	0
031	Phosphorus (%)	2246	0.7100	0.0000	0.7287	0.0265	0.0184	117	-0.70	031.01	0
031	Phosphorus (%)	0682	0.7100	0.0000	0.7287	0.0265	0.0184	117	-0.70	031.43	0
031	Phosphorus (%)	0018	0.7110	0.0120	0.7287	0.0265	0.0184	117	-0.67	031.01	0
031	Phosphorus (%)	2196	0.7130	0.0000	0.7287	0.0265	0.0184	117	-0.59	031.01	0
031	Phosphorus (%)	0407	0.7143	0.0055	0.7287	0.0265	0.0184	117	-0.54	031.43	0
031	Phosphorus (%)	0692	0.7150	0.0500	0.7287	0.0265	0.0184	117	-0.52	031.42	0
031	Phosphorus (%)	0010	0.7150	0.0100	0.7287	0.0265	0.0184	117	-0.52	031.43	0
031	Phosphorus (%)	0425	0.7150	0.0300	0.7287	0.0265	0.0184	117	-0.52	031.43	0
031	Phosphorus (%)	0100	0.7150	0.0100	0.7287	0.0265	0.0184	117	-0.52	031.99	0
031	Phosphorus (%)	2009	0.7160	0.0010	0.7287	0.0265	0.0184	117	-0.48	031.01	0
031	Phosphorus (%)	0693	0.7180	0.0260	0.7287	0.0265	0.0184	117	-0.40	031.42	0
031	Phosphorus (%)	0553	0.7185	0.0950	0.7287	0.0265	0.0184	117	-0.38	031.53	0
031	Phosphorus (%)	0123	0.7200	0.0200	0.7287	0.0265	0.0184	117	-0.33	031.41	0
031	Phosphorus (%)	0171	0.7200	0.0000	0.7287	0.0265	0.0184	117	-0.33	031.41	0
031	Phosphorus (%)	0229	0.7200	0.0000	0.7287	0.0265	0.0184	117	-0.33	031.41	0
031	Phosphorus (%)	0142	0.7203	0.0007	0.7287	0.0265	0.0184	117	-0.32	031.01	0
031	Phosphorus (%)	0186	0.7206	0.0087	0.7287	0.0265	0.0184	117	-0.31	031.52	0
031	Phosphorus (%)	0653	0.7240	0.0020	0.7287	0.0265	0.0184	117	-0.18	031.41	0
031	Phosphorus (%)	0885	0.7243	0.0028	0.7287	0.0265	0.0184	117	-0.16	031.01	0
031	Phosphorus (%)	0563	0.7244	0.0012	0.7287	0.0265	0.0184	117	-0.16	031.01	0
031	Phosphorus (%)	0683	0.7250	0.0100	0.7287	0.0265	0.0184	117	-0.14	031.01	0
031	Phosphorus (%)	2199	0.7250	0.0300	0.7287	0.0265	0.0184	117	-0.14	031.01	0
031	Phosphorus (%)	2302	0.7250	0.0100	0.7287	0.0265	0.0184	117	-0.14	031.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	0910	0.7250	0.0500	0.7287	0.0265	0.0184	117	-0.14	031.41	0
031	Phosphorus (%)	0035	0.7250	0.0100	0.7287	0.0265	0.0184	117	-0.14	031.42	0
031	Phosphorus (%)	0033	0.7250	0.0100	0.7287	0.0265	0.0184	117	-0.14	031.43	0
031	Phosphorus (%)	0876	0.7265	0.0070	0.7287	0.0265	0.0184	117	-0.08	031.43	0
031	Phosphorus (%)	0006	0.7279	0.0074	0.7287	0.0265	0.0184	117	-0.03	031.43	0
031	Phosphorus (%)	0043	0.7285	0.0000	0.7287	0.0265	0.0184	117	-0.01	031.02	0
031	Phosphorus (%)	0884	0.7300	0.0000	0.7287	0.0265	0.0184	117	0.05	031.01	0
031	Phosphorus (%)	2058	0.7300	0.0000	0.7287	0.0265	0.0184	117	0.05	031.01	0
031	Phosphorus (%)	2192	0.7300	0.0000	0.7287	0.0265	0.0184	117	0.05	031.01	0
031	Phosphorus (%)	0144	0.7300	0.0400	0.7287	0.0265	0.0184	117	0.05	031.41	0
031	Phosphorus (%)	0298	0.7300	0.0200	0.7287	0.0265	0.0184	117	0.05	031.41	0
031	Phosphorus (%)	0297	0.7300	0.0200	0.7287	0.0265	0.0184	117	0.05	031.43	0
031	Phosphorus (%)	0001	0.7305	0.0010	0.7287	0.0265	0.0184	117	0.07	031.03	0
031	Phosphorus (%)	0870	0.7311	0.0243	0.7287	0.0265	0.0184	117	0.09	031.42	0
031	Phosphorus (%)	0164	0.7320	0.0220	0.7287	0.0265	0.0184	117	0.13	031.44	0
031	Phosphorus (%)	2006	0.7330	0.0280	0.7287	0.0265	0.0184	117	0.16	031.01	0
031	Phosphorus (%)	0870	0.7331	0.0123	0.7287	0.0265	0.0184	117	0.16	031.43	0
031	Phosphorus (%)	0017	0.7344	0.0169	0.7287	0.0265	0.0184	117	0.21	031.43	0
031	Phosphorus (%)	0169	0.7350	0.0300	0.7287	0.0265	0.0184	117	0.24	031.03	0
031	Phosphorus (%)	0358	0.7350	0.0100	0.7287	0.0265	0.0184	117	0.24	031.41	0
031	Phosphorus (%)	0263	0.7350	0.0000	0.7287	0.0265	0.0184	117	0.24	031.42	0
031	Phosphorus (%)	0939	0.7350	0.0100	0.7287	0.0265	0.0184	117	0.24	031.53	0
031	Phosphorus (%)	0619	0.7370	0.0020	0.7287	0.0265	0.0184	117	0.31	031.41	0
031	Phosphorus (%)	0036	0.7372	0.0145	0.7287	0.0265	0.0184	117	0.32	031.43	0
031	Phosphorus (%)	0881	0.7385	0.0050	0.7287	0.0265	0.0184	117	0.37	031.01	0
031	Phosphorus (%)	2340	0.7395	0.0050	0.7287	0.0265	0.0184	117	0.41	031.01	0
031	Phosphorus (%)	0598	0.7430	0.0022	0.7287	0.0265	0.0184	117	0.54	031.41	0
031	Phosphorus (%)	2190	0.7439	0.0014	0.7287	0.0265	0.0184	117	0.57	031.01	0
031	Phosphorus (%)	2022	0.7450	0.0100	0.7287	0.0265	0.0184	117	0.62	031.01	0
031	Phosphorus (%)	2181	0.7450	0.0500	0.7287	0.0265	0.0184	117	0.62	031.01	0
031	Phosphorus (%)	2325	0.7450	0.0100	0.7287	0.0265	0.0184	117	0.62	031.01	0
031	Phosphorus (%)	0226	0.7450	0.0100	0.7287	0.0265	0.0184	117	0.62	031.41	0
031	Phosphorus (%)	2259	0.7450	0.0260	0.7287	0.0265	0.0184	117	0.62	031.41	0
031	Phosphorus (%)	0675	0.7450	0.0100	0.7287	0.0265	0.0184	117	0.62	031.43	0
031	Phosphorus (%)	2319	0.7460	0.0040	0.7287	0.0265	0.0184	117	0.65	031.42	0
031	Phosphorus (%)	0918	0.7460	0.0160	0.7287	0.0265	0.0184	117	0.65	031.43	0
031	Phosphorus (%)	0357	0.7473	0.0028	0.7287	0.0265	0.0184	117	0.70	031.42	0
031	Phosphorus (%)	0009	0.7474	0.0060	0.7287	0.0265	0.0184	117	0.71	031.42	0
031	Phosphorus (%)	0066	0.7480	0.0180	0.7287	0.0265	0.0184	117	0.73	031.44	0
031	Phosphorus (%)	0042	0.7485	0.0570	0.7287	0.0265	0.0184	117	0.75	031.42	0
031	Phosphorus (%)	0687	0.7500	0.1000	0.7287	0.0265	0.0184	117	0.80	031.01	0
031	Phosphorus (%)	0049	0.7500	0.0000	0.7287	0.0265	0.0184	117	0.80	031.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	0074	0.7500	0.0000	0.7287	0.0265	0.0184	117	0.80	031.41	0
031	Phosphorus (%)	0278	0.7500	0.0400	0.7287	0.0265	0.0184	117	0.80	031.42	0
031	Phosphorus (%)	0505	0.7500	0.0000	0.7287	0.0265	0.0184	117	0.80	031.43	0
031	Phosphorus (%)	0511	0.7500	0.0400	0.7287	0.0265	0.0184	117	0.80	031.43	0
031	Phosphorus (%)	0037	0.7540	0.0100	0.7287	0.0265	0.0184	117	0.95	031.42	0
031	Phosphorus (%)	0186	0.7549	0.0110	0.7287	0.0265	0.0184	117	0.99	031.42	0
031	Phosphorus (%)	0948	0.7550	0.0100	0.7287	0.0265	0.0184	117	0.99	031.01	0
031	Phosphorus (%)	0366	0.7550	0.0300	0.7287	0.0265	0.0184	117	0.99	031.42	0
031	Phosphorus (%)	0035	0.7550	0.0100	0.7287	0.0265	0.0184	117	0.99	031.43	0
031	Phosphorus (%)	2290	0.7561	0.0129	0.7287	0.0265	0.0184	117	1.03	031.99	0
031	Phosphorus (%)	0098	0.7564	0.0081	0.7287	0.0265	0.0184	117	1.04	031.43	0
031	Phosphorus (%)	0968	0.7569	0.0016	0.7287	0.0265	0.0184	117	1.06	031.43	0
031	Phosphorus (%)	0265	0.7600	0.0600	0.7287	0.0265	0.0184	117	1.18	031.42	0
031	Phosphorus (%)	0098	0.7646	0.0241	0.7287	0.0265	0.0184	117	1.35	031.41	0
031	Phosphorus (%)	0345	0.7650	0.0080	0.7287	0.0265	0.0184	117	1.37	031.43	0
031	Phosphorus (%)	0353	0.7700	0.0600	0.7287	0.0265	0.0184	117	1.56	031.43	0
031	Phosphorus (%)	0510	0.7800	0.0200	0.7287	0.0265	0.0184	117	1.93	031.43	0
031	Phosphorus (%)	0148	0.7845	0.0150	0.7287	0.0265	0.0184	117	2.10	031.41	0
031	Phosphorus (%)	0190	0.7950	0.0100	0.7287	0.0265	0.0184	117	2.50	031.42	0
031	Phosphorus (%)	0726	0.8000	0.0000	0.7287	0.0265	0.0184	117	2.69	031.42	0
031	Phosphorus (%)	0520	0.8030	0.0080	0.7287	0.0265	0.0184	117	2.80	031.41	0
031	Phosphorus (%)	0011	1.015	0.1100	0.7287	0.0265	0.0184	117	10.79	031.43	0
031	Phosphorus (%)	0536	1.050	0.1000	0.7287	0.0265	0.0184	117	12.11	031.06	0
031	Phosphorus (%)	0354	1.065	0.0100	0.7287	0.0265	0.0184	117	12.67	031.41	0
031	Phosphorus (%)	0108	0.5600	0.1600	0.7287	0.0265	0.0184	117	-6.36	031.01	1
031	Phosphorus (%)	0047	0.7000	0.1400	0.7287	0.0265	0.0184	117	-1.08	031.42	1
031	Phosphorus (%)	2089	0.9300	0.2000	0.7287	0.0265	0.0184	117	7.59	031.43	1
032	Potassium (%)	0038	0.6305	0.0730	0.8835	0.0571	0.0182	68	-4.43	032.31	0
032	Potassium (%)	0154	0.7768	0.0019	0.8835	0.0571	0.0182	68	-1.87	032.52	0
032	Potassium (%)	2129	0.7875	0.0013	0.8835	0.0571	0.0182	68	-1.68	032.42	0
032	Potassium (%)	0563	0.7908	0.0025	0.8835	0.0571	0.0182	68	-1.62	032.31	0
032	Potassium (%)	2319	0.7950	0.0300	0.8835	0.0571	0.0182	68	-1.55	032.52	0
032	Potassium (%)	0278	0.8000	0.0200	0.8835	0.0571	0.0182	68	-1.46	032.42	0
032	Potassium (%)	0083	0.8020	0.0100	0.8835	0.0571	0.0182	68	-1.43	032.43	0
032	Potassium (%)	0598	0.8029	0.0064	0.8835	0.0571	0.0182	68	-1.41	032.41	0
032	Potassium (%)	0425	0.8200	0.0000	0.8835	0.0571	0.0182	68	-1.11	032.43	0
032	Potassium (%)	0629	0.8300	0.0000	0.8835	0.0571	0.0182	68	-0.94	032.31	0
032	Potassium (%)	0004	0.8300	0.0200	0.8835	0.0571	0.0182	68	-0.94	032.41	0
032	Potassium (%)	0683	0.8350	0.0100	0.8835	0.0571	0.0182	68	-0.85	032.31	0
032	Potassium (%)	0049	0.8350	0.0300	0.8835	0.0571	0.0182	68	-0.85	032.41	0
032	Potassium (%)	0407	0.8394	0.0287	0.8835	0.0571	0.0182	68	-0.77	032.53	0
032	Potassium (%)	0123	0.8400	0.0000	0.8835	0.0571	0.0182	68	-0.76	032.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
032	Potassium (%)	0229	0.8500	0.0200	0.8835	0.0571	0.0182	68	-0.59	032.41	0
032	Potassium (%)	0918	0.8510	0.0180	0.8835	0.0571	0.0182	68	-0.57	032.43	0
032	Potassium (%)	0265	0.8550	0.0700	0.8835	0.0571	0.0182	68	-0.50	032.42	0
032	Potassium (%)	2089	0.8550	0.0100	0.8835	0.0571	0.0182	68	-0.50	032.43	0
032	Potassium (%)	0553	0.8565	0.0190	0.8835	0.0571	0.0182	68	-0.47	032.53	0
032	Potassium (%)	0870	0.8572	0.0070	0.8835	0.0571	0.0182	68	-0.46	032.43	0
032	Potassium (%)	0870	0.8593	0.0148	0.8835	0.0571	0.0182	68	-0.42	032.42	0
032	Potassium (%)	0510	0.8600	0.0000	0.8835	0.0571	0.0182	68	-0.41	032.43	0
032	Potassium (%)	0100	0.8600	0.0200	0.8835	0.0571	0.0182	68	-0.41	032.99	0
032	Potassium (%)	0590	0.8600	0.0800	0.8835	0.0571	0.0182	68	-0.41	032.99	0
032	Potassium (%)	0407	0.8608	0.0077	0.8835	0.0571	0.0182	68	-0.40	032.41	0
032	Potassium (%)	0910	0.8650	0.0300	0.8835	0.0571	0.0182	68	-0.32	032.41	0
032	Potassium (%)	0407	0.8654	0.0074	0.8835	0.0571	0.0182	68	-0.32	032.43	0
032	Potassium (%)	0142	0.8694	0.0044	0.8835	0.0571	0.0182	68	-0.25	032.31	0
032	Potassium (%)	0186	0.8697	0.0040	0.8835	0.0571	0.0182	68	-0.24	032.52	0
032	Potassium (%)	0876	0.8775	0.0350	0.8835	0.0571	0.0182	68	-0.10	032.43	0
032	Potassium (%)	0074	0.8800	0.0200	0.8835	0.0571	0.0182	68	-0.06	032.41	0
032	Potassium (%)	0144	0.8800	0.0600	0.8835	0.0571	0.0182	68	-0.06	032.41	0
032	Potassium (%)	0098	0.8850	0.0100	0.8835	0.0571	0.0182	68	0.03	032.43	0
032	Potassium (%)	0164	0.8885	0.0270	0.8835	0.0571	0.0182	68	0.09	032.44	0
032	Potassium (%)	0682	0.8900	0.0000	0.8835	0.0571	0.0182	68	0.11	032.43	0
032	Potassium (%)	0045	0.8915	0.0250	0.8835	0.0571	0.0182	68	0.14	032.42	0
032	Potassium (%)	0572	0.8920	0.0120	0.8835	0.0571	0.0182	68	0.15	032.53	0
032	Potassium (%)	0006	0.8924	0.0076	0.8835	0.0571	0.0182	68	0.16	032.43	0
032	Potassium (%)	0692	0.8950	0.0300	0.8835	0.0571	0.0182	68	0.20	032.42	0
032	Potassium (%)	0035	0.8950	0.0100	0.8835	0.0571	0.0182	68	0.20	032.43	0
032	Potassium (%)	0968	0.8952	0.0002	0.8835	0.0571	0.0182	68	0.20	032.43	0
032	Potassium (%)	0042	0.8965	0.0150	0.8835	0.0571	0.0182	68	0.23	032.42	0
032	Potassium (%)	0098	0.9000	0.0400	0.8835	0.0571	0.0182	68	0.29	032.41	0
032	Potassium (%)	0035	0.9000	0.0000	0.8835	0.0571	0.0182	68	0.29	032.42	0
032	Potassium (%)	0357	0.9050	0.0025	0.8835	0.0571	0.0182	68	0.38	032.42	0
032	Potassium (%)	0171	0.9050	0.0100	0.8835	0.0571	0.0182	68	0.38	032.41	0
032	Potassium (%)	0358	0.9050	0.0100	0.8835	0.0571	0.0182	68	0.38	032.41	0
032	Potassium (%)	0939	0.9050	0.0100	0.8835	0.0571	0.0182	68	0.38	032.53	0
032	Potassium (%)	0037	0.9070	0.0040	0.8835	0.0571	0.0182	68	0.41	032.42	0
032	Potassium (%)	0675	0.9150	0.0100	0.8835	0.0571	0.0182	68	0.55	032.43	0
032	Potassium (%)	0297	0.9250	0.0300	0.8835	0.0571	0.0182	68	0.73	032.43	0
032	Potassium (%)	0345	0.9310	0.0080	0.8835	0.0571	0.0182	68	0.83	032.43	0
032	Potassium (%)	0520	0.9315	0.0070	0.8835	0.0571	0.0182	68	0.84	032.41	0
032	Potassium (%)	0889	0.9350	0.0100	0.8835	0.0571	0.0182	68	0.90	032.99	0
032	Potassium (%)	0263	0.9360	0.0100	0.8835	0.0571	0.0182	68	0.92	032.42	0
032	Potassium (%)	0148	0.9380	0.0080	0.8835	0.0571	0.0182	68	0.95	032.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
032	Potassium (%)	0226	0.9400	0.0000	0.8835	0.0571	0.0182	68	0.99	032.41	0
032	Potassium (%)	0186	0.9444	0.0224	0.8835	0.0571	0.0182	68	1.07	032.42	0
032	Potassium (%)	0353	0.9450	0.0100	0.8835	0.0571	0.0182	68	1.08	032.43	0
032	Potassium (%)	0653	0.9490	0.0060	0.8835	0.0571	0.0182	68	1.15	032.41	0
032	Potassium (%)	0693	0.9720	0.0220	0.8835	0.0571	0.0182	68	1.55	032.42	0
032	Potassium (%)	0726	0.9800	0.0000	0.8835	0.0571	0.0182	68	1.69	032.42	0
032	Potassium (%)	2188	0.9900	0.0000	0.8835	0.0571	0.0182	68	1.86	032.08	0
032	Potassium (%)	0169	1.055	0.0500	0.8835	0.0571	0.0182	68	3.00	032.32	0
032	Potassium (%)	0366	1.070	0.0400	0.8835	0.0571	0.0182	68	3.26	032.42	0
032	Potassium (%)	0536	1.250	0.1000	0.8835	0.0571	0.0182	68	6.41	032.31	0
032	Potassium (%)	0354	1.295	0.0300	0.8835	0.0571	0.0182	68	7.20	032.41	0
032	Potassium (%)	0047	0.7400	0.2000	0.8835	0.0571	0.0182	68	-2.51	032.42	1
032	Potassium (%)	2302	0.9500	0.1600	0.8835	0.0571	0.0182	68	1.16	032.99	1
033	Salt as chloride (%)	0366	0.4850	0.0300	0.7567	0.0422	0.0173	70	-6.43	033.00	0
033	Salt as chloride (%)	0505	0.4950	0.0300	0.7567	0.0422	0.0173	70	-6.20	033.03	0
033	Salt as chloride (%)	0265	0.6300	0.0600	0.7567	0.0422	0.0173	70	-3.00	033.03	0
033	Salt as chloride (%)	0171	0.6350	0.0100	0.7567	0.0422	0.0173	70	-2.88	033.05	0
033	Salt as chloride (%)	0619	0.6410	0.0020	0.7567	0.0422	0.0173	70	-2.74	033.01	0
033	Salt as chloride (%)	0511	0.6550	0.0300	0.7567	0.0422	0.0173	70	-2.41	033.00	0
033	Salt as chloride (%)	2246	0.6550	0.0000	0.7567	0.0422	0.0173	70	-2.41	033.99	0
033	Salt as chloride (%)	2337	0.6800	0.0200	0.7567	0.0422	0.0173	70	-1.82	033.00	0
033	Salt as chloride (%)	2009	0.6932	0.0056	0.7567	0.0422	0.0173	70	-1.50	033.00	0
033	Salt as chloride (%)	0358	0.7000	0.0000	0.7567	0.0422	0.0173	70	-1.34	033.99	0
033	Salt as chloride (%)	0123	0.7050	0.0100	0.7567	0.0422	0.0173	70	-1.23	033.00	0
033	Salt as chloride (%)	0017	0.7097	0.0248	0.7567	0.0422	0.0173	70	-1.11	033.99	0
033	Salt as chloride (%)	0510	0.7175	0.0170	0.7567	0.0422	0.0173	70	-0.93	033.01	0
033	Salt as chloride (%)	2196	0.7200	0.0000	0.7567	0.0422	0.0173	70	-0.87	033.00	0
033	Salt as chloride (%)	2058	0.7200	0.0400	0.7567	0.0422	0.0173	70	-0.87	033.99	0
033	Salt as chloride (%)	0653	0.7260	0.0100	0.7567	0.0422	0.0173	70	-0.73	033.00	0
033	Salt as chloride (%)	0910	0.7300	0.0000	0.7567	0.0422	0.0173	70	-0.63	033.99	0
033	Salt as chloride (%)	0689	0.7350	0.0300	0.7567	0.0422	0.0173	70	-0.51	033.05	0
033	Salt as chloride (%)	0038	0.7370	0.1560	0.7567	0.0422	0.0173	70	-0.47	033.01	0
033	Salt as chloride (%)	0674	0.7400	0.0000	0.7567	0.0422	0.0173	70	-0.40	033.00	0
033	Salt as chloride (%)	0425	0.7400	0.0000	0.7567	0.0422	0.0173	70	-0.40	033.01	0
033	Salt as chloride (%)	0726	0.7400	0.0000	0.7567	0.0422	0.0173	70	-0.40	033.03	0
033	Salt as chloride (%)	0309	0.7432	0.0001	0.7567	0.0422	0.0173	70	-0.32	033.00	0
033	Salt as chloride (%)	2109	0.7475	0.0110	0.7567	0.0422	0.0173	70	-0.22	033.01	0
033	Salt as chloride (%)	0885	0.7479	0.0107	0.7567	0.0422	0.0173	70	-0.21	033.00	0
033	Salt as chloride (%)	0693	0.7500	0.0160	0.7567	0.0422	0.0173	70	-0.16	033.00	0
033	Salt as chloride (%)	2193	0.7500	0.0200	0.7567	0.0422	0.0173	70	-0.16	033.00	0
033	Salt as chloride (%)	0003	0.7500	0.1000	0.7567	0.0422	0.0173	70	-0.16	033.01	0
033	Salt as chloride (%)	0675	0.7550	0.0100	0.7567	0.0422	0.0173	70	-0.04	033.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
033	Salt as chloride (%)	0278	0.7550	0.0300	0.7567	0.0422	0.0173	70	-0.04	033.01	0
033	Salt as chloride (%)	0881	0.7555	0.0050	0.7567	0.0422	0.0173	70	-0.03	033.00	0
033	Salt as chloride (%)	0148	0.7600	0.0140	0.7567	0.0422	0.0173	70	0.08	033.01	0
033	Salt as chloride (%)	0190	0.7600	0.0000	0.7567	0.0422	0.0173	70	0.08	033.03	0
033	Salt as chloride (%)	2146	0.7650	0.0100	0.7567	0.0422	0.0173	70	0.20	033.00	0
033	Salt as chloride (%)	0001	0.7650	0.0100	0.7567	0.0422	0.0173	70	0.20	033.01	0
033	Salt as chloride (%)	0004	0.7650	0.0100	0.7567	0.0422	0.0173	70	0.20	033.01	0
033	Salt as chloride (%)	0028	0.7650	0.0100	0.7567	0.0422	0.0173	70	0.20	033.01	0
033	Salt as chloride (%)	0083	0.7650	0.0100	0.7567	0.0422	0.0173	70	0.20	033.01	0
033	Salt as chloride (%)	0098	0.7650	0.0100	0.7567	0.0422	0.0173	70	0.20	033.01	0
033	Salt as chloride (%)	0590	0.7650	0.0100	0.7567	0.0422	0.0173	70	0.20	033.01	0
033	Salt as chloride (%)	0683	0.7650	0.0100	0.7567	0.0422	0.0173	70	0.20	033.01	0
033	Salt as chloride (%)	0939	0.7650	0.0100	0.7567	0.0422	0.0173	70	0.20	033.01	0
033	Salt as chloride (%)	2319	0.7650	0.0100	0.7567	0.0422	0.0173	70	0.20	033.01	0
033	Salt as chloride (%)	0650	0.7650	0.0100	0.7567	0.0422	0.0173	70	0.20	033.99	0
033	Salt as chloride (%)	0354	0.7665	0.0070	0.7567	0.0422	0.0173	70	0.23	033.01	0
033	Salt as chloride (%)	0229	0.7700	0.0000	0.7567	0.0422	0.0173	70	0.31	033.01	0
033	Salt as chloride (%)	2190	0.7710	0.0080	0.7567	0.0422	0.0173	70	0.34	033.00	0
033	Salt as chloride (%)	0164	0.7750	0.0100	0.7567	0.0422	0.0173	70	0.43	033.01	0
033	Salt as chloride (%)	0407	0.7780	0.0160	0.7567	0.0422	0.0173	70	0.50	033.01	0
033	Salt as chloride (%)	0298	0.7800	0.0200	0.7567	0.0422	0.0173	70	0.55	033.00	0
033	Salt as chloride (%)	0100	0.7800	0.0000	0.7567	0.0422	0.0173	70	0.55	033.01	0
033	Salt as chloride (%)	2146	0.7800	0.0000	0.7567	0.0422	0.0173	70	0.55	033.01	0
033	Salt as chloride (%)	0629	0.7800	0.0200	0.7567	0.0422	0.0173	70	0.55	033.99	0
033	Salt as chloride (%)	2181	0.7800	0.0600	0.7567	0.0422	0.0173	70	0.55	033.99	0
033	Salt as chloride (%)	0045	0.7805	0.0210	0.7567	0.0422	0.0173	70	0.56	033.00	0
033	Salt as chloride (%)	0948	0.7850	0.0100	0.7567	0.0422	0.0173	70	0.67	033.01	0
033	Salt as chloride (%)	2129	0.7895	0.0370	0.7567	0.0422	0.0173	70	0.78	033.99	0
033	Salt as chloride (%)	0353	0.7900	0.0400	0.7567	0.0422	0.0173	70	0.79	033.00	0
033	Salt as chloride (%)	0019	0.7900	0.0200	0.7567	0.0422	0.0173	70	0.79	033.01	0
033	Salt as chloride (%)	0918	0.7980	0.0000	0.7567	0.0422	0.0173	70	0.98	033.00	0
033	Salt as chloride (%)	2268	0.8000	0.0200	0.7567	0.0422	0.0173	70	1.02	033.01	0
033	Salt as chloride (%)	0914	0.8000	0.0000	0.7567	0.0422	0.0173	70	1.02	033.99	0
033	Salt as chloride (%)	0186	0.8030	0.0300	0.7567	0.0422	0.0173	70	1.10	033.01	0
033	Salt as chloride (%)	0226	0.8050	0.0100	0.7567	0.0422	0.0173	70	1.14	033.01	0
033	Salt as chloride (%)	2076	0.8390	0.0280	0.7567	0.0422	0.0173	70	1.95	033.00	0
033	Salt as chloride (%)	0539	0.8400	0.0400	0.7567	0.0422	0.0173	70	1.97	033.00	0
033	Salt as chloride (%)	0682	0.8400	0.0000	0.7567	0.0422	0.0173	70	1.97	033.01	0
033	Salt as chloride (%)	2188	0.8500	0.0000	0.7567	0.0422	0.0173	70	2.21	033.05	0
033	Salt as chloride (%)	0144	0.8800	0.0000	0.7567	0.0422	0.0173	70	2.92	033.03	0
033	Salt as chloride (%)	2073	0.9550	0.0100	0.7567	0.0422	0.0173	70	4.69	033.99	0
033	Salt as chloride (%)	0536	1.300	0.4000	0.7567	0.0422	0.0173	70	12.86	033.99	2

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
034	Selenium, Total (Se) (ppm)	0045	0.5835	0.1350	0.6840	0.1009	0.0510	14	-1.00	034.04	0
034	Selenium, Total (Se) (ppm)	0164	0.5945	0.0150	0.6840	0.1009	0.0510	14	-0.89	034.53	0
034	Selenium, Total (Se) (ppm)	0171	0.6005	0.0010	0.6840	0.1009	0.0510	14	-0.83	034.04	0
034	Selenium, Total (Se) (ppm)	0619	0.6145	0.0110	0.6840	0.1009	0.0510	14	-0.69	034.41	0
034	Selenium, Total (Se) (ppm)	0910	0.6250	0.0300	0.6840	0.1009	0.0510	14	-0.58	034.52	0
034	Selenium, Total (Se) (ppm)	0553	0.6250	0.0260	0.6840	0.1009	0.0510	14	-0.58	034.53	0
034	Selenium, Total (Se) (ppm)	0169	0.6300	0.0200	0.6840	0.1009	0.0510	14	-0.53	034.04	0
034	Selenium, Total (Se) (ppm)	0560	0.6650	0.0300	0.6840	0.1009	0.0510	14	-0.19	034.53	0
034	Selenium, Total (Se) (ppm)	0407	0.7095	0.0942	0.6840	0.1009	0.0510	14	0.25	034.53	0
034	Selenium, Total (Se) (ppm)	0038	0.7180	0.0920	0.6840	0.1009	0.0510	14	0.34	034.01	0
034	Selenium, Total (Se) (ppm)	0563	0.7446	0.0297	0.6840	0.1009	0.0510	14	0.60	034.04	0
034	Selenium, Total (Se) (ppm)	0939	0.8100	0.0800	0.6840	0.1009	0.0510	14	1.25	034.53	0
034	Selenium, Total (Se) (ppm)	0186	0.8200	0.1000	0.6840	0.1009	0.0510	14	1.35	034.52	0
034	Selenium, Total (Se) (ppm)	0190	1.205	0.0500	0.6840	0.1009	0.0510	14	5.17	034.99	0
034	Selenium, Total (Se) (ppm)	0692	1.400	0.6000	0.6840	0.1009	0.0510	14	7.10	034.42	1
034	Selenium, Total (Se) (ppm)	0682	< 5		0.6840	0.1009	0.0510	14		034.43	5
035	Sodium (%)	0889	0.0750	0.0100	0.2710	0.0145	0.0078	72	-13.48	035.99	0
035	Sodium (%)	0536	0.2350	0.0100	0.2710	0.0145	0.0078	72	-2.47	035.31	0
035	Sodium (%)	2129	0.2450	0.0053	0.2710	0.0145	0.0078	72	-1.79	035.42	0
035	Sodium (%)	0683	0.2450	0.0100	0.2710	0.0145	0.0078	72	-1.79	035.31	0
035	Sodium (%)	0590	0.2450	0.0100	0.2710	0.0145	0.0078	72	-1.79	035.99	0
035	Sodium (%)	0300	0.2475	0.0350	0.2710	0.0145	0.0078	72	-1.61	035.33	0
035	Sodium (%)	0629	0.2500	0.0000	0.2710	0.0145	0.0078	72	-1.44	035.31	0
035	Sodium (%)	0407	0.2503	0.0108	0.2710	0.0145	0.0078	72	-1.42	035.53	0
035	Sodium (%)	0563	0.2543	0.0078	0.2710	0.0145	0.0078	72	-1.15	035.31	0
035	Sodium (%)	0265	0.2550	0.0300	0.2710	0.0145	0.0078	72	-1.10	035.42	0
035	Sodium (%)	0366	0.2550	0.0160	0.2710	0.0145	0.0078	72	-1.10	035.42	0
035	Sodium (%)	0045	0.2560	0.0180	0.2710	0.0145	0.0078	72	-1.03	035.42	0
035	Sodium (%)	0407	0.2571	0.0094	0.2710	0.0145	0.0078	72	-0.95	035.41	0
035	Sodium (%)	0298	0.2600	0.0200	0.2710	0.0145	0.0078	72	-0.75	035.41	0
035	Sodium (%)	2089	0.2600	0.0000	0.2710	0.0145	0.0078	72	-0.75	035.43	0
035	Sodium (%)	0939	0.2600	0.0000	0.2710	0.0145	0.0078	72	-0.75	035.53	0
035	Sodium (%)	0407	0.2630	0.0008	0.2710	0.0145	0.0078	72	-0.55	035.43	0
035	Sodium (%)	0510	0.2635	0.0010	0.2710	0.0145	0.0078	72	-0.51	035.43	0
035	Sodium (%)	0263	0.2645	0.0010	0.2710	0.0145	0.0078	72	-0.45	035.42	0
035	Sodium (%)	0870	0.2645	0.0082	0.2710	0.0145	0.0078	72	-0.45	035.42	0
035	Sodium (%)	0918	0.2645	0.0010	0.2710	0.0145	0.0078	72	-0.45	035.43	0
035	Sodium (%)	2246	0.2650	0.0100	0.2710	0.0145	0.0078	72	-0.41	035.31	0
035	Sodium (%)	0049	0.2650	0.0100	0.2710	0.0145	0.0078	72	-0.41	035.41	0
035	Sodium (%)	0229	0.2650	0.0100	0.2710	0.0145	0.0078	72	-0.41	035.41	0
035	Sodium (%)	0619	0.2650	0.0020	0.2710	0.0145	0.0078	72	-0.41	035.41	0
035	Sodium (%)	0278	0.2650	0.0100	0.2710	0.0145	0.0078	72	-0.41	035.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
035	Sodium (%)	0675	0.2650	0.0100	0.2710	0.0145	0.0078	72	-0.41	035.43	0
035	Sodium (%)	0100	0.2650	0.0100	0.2710	0.0145	0.0078	72	-0.41	035.99	0
035	Sodium (%)	0870	0.2654	0.0072	0.2710	0.0145	0.0078	72	-0.38	035.43	0
035	Sodium (%)	0876	0.2655	0.0150	0.2710	0.0145	0.0078	72	-0.38	035.43	0
035	Sodium (%)	0357	0.2674	0.0026	0.2710	0.0145	0.0078	72	-0.25	035.42	0
035	Sodium (%)	0154	0.2677	0.0009	0.2710	0.0145	0.0078	72	-0.23	035.52	0
035	Sodium (%)	2022	0.2700	0.0080	0.2710	0.0145	0.0078	72	-0.07	035.31	0
035	Sodium (%)	0144	0.2700	0.0000	0.2710	0.0145	0.0078	72	-0.07	035.41	0
035	Sodium (%)	0726	0.2700	0.0000	0.2710	0.0145	0.0078	72	-0.07	035.42	0
035	Sodium (%)	0682	0.2700	0.0000	0.2710	0.0145	0.0078	72	-0.07	035.43	0
035	Sodium (%)	2319	0.2715	0.0010	0.2710	0.0145	0.0078	72	0.04	035.42	0
035	Sodium (%)	0083	0.2730	0.0000	0.2710	0.0145	0.0078	72	0.14	035.43	0
035	Sodium (%)	0553	0.2730	0.0020	0.2710	0.0145	0.0078	72	0.14	035.53	0
035	Sodium (%)	0037	0.2740	0.0120	0.2710	0.0145	0.0078	72	0.21	035.42	0
035	Sodium (%)	0345	0.2745	0.0030	0.2710	0.0145	0.0078	72	0.24	035.43	0
035	Sodium (%)	0108	0.2750	0.0100	0.2710	0.0145	0.0078	72	0.28	035.05	0
035	Sodium (%)	0004	0.2750	0.0100	0.2710	0.0145	0.0078	72	0.28	035.41	0
035	Sodium (%)	0098	0.2750	0.0100	0.2710	0.0145	0.0078	72	0.28	035.41	0
035	Sodium (%)	0123	0.2750	0.0100	0.2710	0.0145	0.0078	72	0.28	035.41	0
035	Sodium (%)	0164	0.2750	0.0080	0.2710	0.0145	0.0078	72	0.28	035.41	0
035	Sodium (%)	0098	0.2750	0.0100	0.2710	0.0145	0.0078	72	0.28	035.43	0
035	Sodium (%)	0598	0.2752	0.0009	0.2710	0.0145	0.0078	72	0.29	035.41	0
035	Sodium (%)	0006	0.2765	0.0033	0.2710	0.0145	0.0078	72	0.38	035.43	0
035	Sodium (%)	0142	0.2776	0.0069	0.2710	0.0145	0.0078	72	0.45	035.31	0
035	Sodium (%)	2188	0.2800	0.0000	0.2710	0.0145	0.0078	72	0.62	035.01	0
035	Sodium (%)	2196	0.2800	0.0000	0.2710	0.0145	0.0078	72	0.62	035.31	0
035	Sodium (%)	0171	0.2800	0.0000	0.2710	0.0145	0.0078	72	0.62	035.41	0
035	Sodium (%)	0358	0.2800	0.0000	0.2710	0.0145	0.0078	72	0.62	035.41	0
035	Sodium (%)	0910	0.2800	0.0000	0.2710	0.0145	0.0078	72	0.62	035.41	0
035	Sodium (%)	0035	0.2800	0.0000	0.2710	0.0145	0.0078	72	0.62	035.42	0
035	Sodium (%)	0035	0.2800	0.0000	0.2710	0.0145	0.0078	72	0.62	035.43	0
035	Sodium (%)	0353	0.2802	0.0031	0.2710	0.0145	0.0078	72	0.63	035.43	0
035	Sodium (%)	0572	0.2815	0.0090	0.2710	0.0145	0.0078	72	0.72	035.53	0
035	Sodium (%)	0693	0.2840	0.0120	0.2710	0.0145	0.0078	72	0.90	035.42	0
035	Sodium (%)	0186	0.2842	0.0026	0.2710	0.0145	0.0078	72	0.91	035.52	0
035	Sodium (%)	0297	0.2850	0.0100	0.2710	0.0145	0.0078	72	0.97	035.43	0
035	Sodium (%)	0042	0.2895	0.0330	0.2710	0.0145	0.0078	72	1.27	035.42	0
035	Sodium (%)	0038	0.2900	0.0040	0.2710	0.0145	0.0078	72	1.31	035.31	0
035	Sodium (%)	0148	0.2900	0.0100	0.2710	0.0145	0.0078	72	1.31	035.41	0
035	Sodium (%)	0692	0.2900	0.0400	0.2710	0.0145	0.0078	72	1.31	035.42	0
035	Sodium (%)	0653	0.2960	0.0040	0.2710	0.0145	0.0078	72	1.72	035.41	0
035	Sodium (%)	0186	0.3029	0.0062	0.2710	0.0145	0.0078	72	2.20	035.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
035	Sodium (%)	0520	0.3107	0.0187	0.2710	0.0145	0.0078	72	2.73	035.41	0
035	Sodium (%)	0226	0.3150	0.0100	0.2710	0.0145	0.0078	72	3.03	035.41	0
035	Sodium (%)	0425	0.3150	0.0100	0.2710	0.0145	0.0078	72	3.03	035.43	0
035	Sodium (%)	2109	0.3225	0.0050	0.2710	0.0145	0.0078	72	3.54	035.01	0
035	Sodium (%)	0169	0.3300	0.1000	0.2710	0.0145	0.0078	72	4.06	035.32	1
035	Sodium (%)	2302	0.3500	0.0800	0.2710	0.0145	0.0078	72	5.44	035.99	1
035	Sodium (%)	0354	0.5200	0.0000	0.2710	0.0145	0.0078	72	17.13	035.41	2
036	Sulfur (%)	0229	0.2200	0.0000	0.2881	0.0216	0.0100	42	-3.16	036.42	0
036	Sulfur (%)	0083	0.2350	0.0020	0.2881	0.0216	0.0100	42	-2.46	036.43	0
036	Sulfur (%)	2129	0.2504	0.0137	0.2881	0.0216	0.0100	42	-1.75	036.42	0
036	Sulfur (%)	0692	0.2550	0.0100	0.2881	0.0216	0.0100	42	-1.53	036.42	0
036	Sulfur (%)	0045	0.2600	0.0020	0.2881	0.0216	0.0100	42	-1.30	036.42	0
036	Sulfur (%)	0870	0.2617	0.0060	0.2881	0.0216	0.0100	42	-1.22	036.42	0
036	Sulfur (%)	0366	0.2700	0.0200	0.2881	0.0216	0.0100	42	-0.84	036.42	0
036	Sulfur (%)	0297	0.2700	0.0400	0.2881	0.0216	0.0100	42	-0.84	036.43	0
036	Sulfur (%)	0889	0.2700	0.0000	0.2881	0.0216	0.0100	42	-0.84	036.99	0
036	Sulfur (%)	0407	0.2713	0.0201	0.2881	0.0216	0.0100	42	-0.78	036.42	0
036	Sulfur (%)	0870	0.2741	0.0081	0.2881	0.0216	0.0100	42	-0.65	036.43	0
036	Sulfur (%)	0553	0.2785	0.0290	0.2881	0.0216	0.0100	42	-0.44	036.53	0
036	Sulfur (%)	0357	0.2789	0.0008	0.2881	0.0216	0.0100	42	-0.42	036.42	0
036	Sulfur (%)	0171	0.2800	0.0000	0.2881	0.0216	0.0100	42	-0.37	036.42	0
036	Sulfur (%)	0042	0.2815	0.0030	0.2881	0.0216	0.0100	42	-0.30	036.42	0
036	Sulfur (%)	0407	0.2819	0.0079	0.2881	0.0216	0.0100	42	-0.29	036.43	0
036	Sulfur (%)	0148	0.2830	0.0100	0.2881	0.0216	0.0100	42	-0.23	036.04	0
036	Sulfur (%)	0229	0.2850	0.0100	0.2881	0.0216	0.0100	42	-0.14	036.04	0
036	Sulfur (%)	0035	0.2850	0.0100	0.2881	0.0216	0.0100	42	-0.14	036.42	0
036	Sulfur (%)	0265	0.2850	0.0100	0.2881	0.0216	0.0100	42	-0.14	036.42	0
036	Sulfur (%)	0035	0.2850	0.0100	0.2881	0.0216	0.0100	42	-0.14	036.43	0
036	Sulfur (%)	0693	0.2880	0.0100	0.2881	0.0216	0.0100	42	0.00	036.42	0
036	Sulfur (%)	0278	0.2900	0.0000	0.2881	0.0216	0.0100	42	0.09	036.42	0
036	Sulfur (%)	0033	0.2900	0.0000	0.2881	0.0216	0.0100	42	0.09	036.43	0
036	Sulfur (%)	0708	0.2935	0.0030	0.2881	0.0216	0.0100	42	0.25	036.42	0
036	Sulfur (%)	0876	0.2945	0.0030	0.2881	0.0216	0.0100	42	0.30	036.43	0
036	Sulfur (%)	0164	0.2950	0.0100	0.2881	0.0216	0.0100	42	0.32	036.42	0
036	Sulfur (%)	0098	0.2985	0.0075	0.2881	0.0216	0.0100	42	0.48	036.43	0
036	Sulfur (%)	0186	0.2992	0.0005	0.2881	0.0216	0.0100	42	0.51	036.42	0
036	Sulfur (%)	0226	0.3000	0.0000	0.2881	0.0216	0.0100	42	0.55	036.04	0
036	Sulfur (%)	0726	0.3000	0.0000	0.2881	0.0216	0.0100	42	0.55	036.42	0
036	Sulfur (%)	0353	0.3000	0.0000	0.2881	0.0216	0.0100	42	0.55	036.43	0
036	Sulfur (%)	0918	0.3025	0.0210	0.2881	0.0216	0.0100	42	0.67	036.43	0
036	Sulfur (%)	0345	0.3035	0.0050	0.2881	0.0216	0.0100	42	0.72	036.43	0
036	Sulfur (%)	0169	0.3050	0.0100	0.2881	0.0216	0.0100	42	0.79	036.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
036	Sulfur (%)	0186	0.3088	0.0077	0.2881	0.0216	0.0100	42	0.96	036.52	0
036	Sulfur (%)	0682	0.3100	0.0000	0.2881	0.0216	0.0100	42	1.02	036.43	0
036	Sulfur (%)	0190	0.3150	0.0100	0.2881	0.0216	0.0100	42	1.25	036.42	0
036	Sulfur (%)	2089	0.3200	0.0800	0.2881	0.0216	0.0100	42	1.48	036.43	0
036	Sulfur (%)	0037	0.3265	0.0190	0.2881	0.0216	0.0100	42	1.78	036.42	0
036	Sulfur (%)	0505	0.3350	0.0100	0.2881	0.0216	0.0100	42	2.18	036.43	0
036	Sulfur (%)	0510	0.3350	0.0100	0.2881	0.0216	0.0100	42	2.18	036.43	0
036	Sulfur (%)	0910	2,336	11.00	0.2881	0.0216	0.0100	42	108312.09	036.42	2
037	Zinc (ppm)	0047	105.0	2.270	138.9	9.978	3.804	79	-3.39	037.52	0
037	Zinc (ppm)	0407	112.8	5.887	138.9	9.978	3.804	79	-2.61	037.53	0
037	Zinc (ppm)	0590	117.0	2.000	138.9	9.978	3.804	79	-2.19	037.99	0
037	Zinc (ppm)	0045	124.0	2.000	138.9	9.978	3.804	79	-1.49	037.42	0
037	Zinc (ppm)	0939	124.0	0.0000	138.9	9.978	3.804	79	-1.49	037.53	0
037	Zinc (ppm)	0511	125.0	0.0000	138.9	9.978	3.804	79	-1.39	037.43	0
037	Zinc (ppm)	0009	125.5	3.000	138.9	9.978	3.804	79	-1.34	037.42	0
037	Zinc (ppm)	2129	125.5	1.400	138.9	9.978	3.804	79	-1.34	037.42	0
037	Zinc (ppm)	0164	125.5	1.000	138.9	9.978	3.804	79	-1.34	037.44	0
037	Zinc (ppm)	0366	126.0	2.000	138.9	9.978	3.804	79	-1.29	037.42	0
037	Zinc (ppm)	0353	126.0	4.000	138.9	9.978	3.804	79	-1.29	037.43	0
037	Zinc (ppm)	0049	127.2	4.030	138.9	9.978	3.804	79	-1.17	037.41	0
037	Zinc (ppm)	0358	129.3	2.000	138.9	9.978	3.804	79	-0.96	037.41	0
037	Zinc (ppm)	0870	129.3	1.200	138.9	9.978	3.804	79	-0.96	037.42	0
037	Zinc (ppm)	0889	129.7	2.740	138.9	9.978	3.804	79	-0.92	037.99	0
037	Zinc (ppm)	0619	130.8	3.093	138.9	9.978	3.804	79	-0.80	037.41	0
037	Zinc (ppm)	0017	130.9	8.380	138.9	9.978	3.804	79	-0.80	037.43	0
037	Zinc (ppm)	0692	131.4	18.75	138.9	9.978	3.804	79	-0.75	037.42	0
037	Zinc (ppm)	0263	131.4	1.147	138.9	9.978	3.804	79	-0.75	037.42	0
037	Zinc (ppm)	0918	131.5	0.8400	138.9	9.978	3.804	79	-0.74	037.43	0
037	Zinc (ppm)	0529	132.9	2.200	138.9	9.978	3.804	79	-0.60	037.31	0
037	Zinc (ppm)	0870	133.2	0.9000	138.9	9.978	3.804	79	-0.57	037.43	0
037	Zinc (ppm)	0278	134.4	1.400	138.9	9.978	3.804	79	-0.45	037.42	0
037	Zinc (ppm)	0083	134.5	1.000	138.9	9.978	3.804	79	-0.44	037.43	0
037	Zinc (ppm)	0689	135.0	0.0000	138.9	9.978	3.804	79	-0.39	037.31	0
037	Zinc (ppm)	0074	135.0	14.00	138.9	9.978	3.804	79	-0.39	037.41	0
037	Zinc (ppm)	0876	135.0	16.00	138.9	9.978	3.804	79	-0.39	037.43	0
037	Zinc (ppm)	0572	135.5	3.000	138.9	9.978	3.804	79	-0.34	037.53	0
037	Zinc (ppm)	0563	136.0	0.6083	138.9	9.978	3.804	79	-0.29	037.31	0
037	Zinc (ppm)	0066	136.3	9.570	138.9	9.978	3.804	79	-0.26	037.44	0
037	Zinc (ppm)	0019	136.6	9.070	138.9	9.978	3.804	79	-0.23	037.41	0
037	Zinc (ppm)	0651	136.6	1.175	138.9	9.978	3.804	79	-0.22	037.31	0
037	Zinc (ppm)	0520	136.7	2.720	138.9	9.978	3.804	79	-0.22	037.41	0
037	Zinc (ppm)	0186	137.0	5.000	138.9	9.978	3.804	79	-0.19	037.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
037	Zinc (ppm)	2319	137.5	3.000	138.9	9.978	3.804	79	-0.14	037.42	0
037	Zinc (ppm)	0425	137.6	1.000	138.9	9.978	3.804	79	-0.13	037.43	0
037	Zinc (ppm)	2089	137.8	4.940	138.9	9.978	3.804	79	-0.11	037.43	0
037	Zinc (ppm)	0848	138.2	1.970	138.9	9.978	3.804	79	-0.06	037.41	0
037	Zinc (ppm)	0598	138.5	1.208	138.9	9.978	3.804	79	-0.04	037.41	0
037	Zinc (ppm)	0169	139.0	2.000	138.9	9.978	3.804	79	0.01	037.43	0
037	Zinc (ppm)	0674	139.6	0.0000	138.9	9.978	3.804	79	0.07	037.31	0
037	Zinc (ppm)	0682	140.0	0.0000	138.9	9.978	3.804	79	0.11	037.43	0
037	Zinc (ppm)	2022	140.5	3.000	138.9	9.978	3.804	79	0.16	037.31	0
037	Zinc (ppm)	0027	140.5	3.704	138.9	9.978	3.804	79	0.17	037.43	0
037	Zinc (ppm)	0653	141.2	0.8000	138.9	9.978	3.804	79	0.23	037.41	0
037	Zinc (ppm)	0001	141.5	7.500	138.9	9.978	3.804	79	0.26	037.31	0
037	Zinc (ppm)	0553	141.5	5.000	138.9	9.978	3.804	79	0.26	037.53	0
037	Zinc (ppm)	0226	142.0	1.640	138.9	9.978	3.804	79	0.31	037.41	0
037	Zinc (ppm)	0043	142.1	5.213	138.9	9.978	3.804	79	0.32	037.43	0
037	Zinc (ppm)	0345	142.2	4.100	138.9	9.978	3.804	79	0.33	037.43	0
037	Zinc (ppm)	0407	142.3	6.922	138.9	9.978	3.804	79	0.35	037.43	0
037	Zinc (ppm)	2196	143.0	0.0000	138.9	9.978	3.804	79	0.41	037.31	0
037	Zinc (ppm)	0510	143.0	2.000	138.9	9.978	3.804	79	0.42	037.43	0
037	Zinc (ppm)	0171	143.5	0.6000	138.9	9.978	3.804	79	0.47	037.34	0
037	Zinc (ppm)	0910	143.5	9.000	138.9	9.978	3.804	79	0.47	037.41	0
037	Zinc (ppm)	0693	143.7	14.02	138.9	9.978	3.804	79	0.49	037.42	0
037	Zinc (ppm)	2246	144.0	3.580	138.9	9.978	3.804	79	0.52	037.31	0
037	Zinc (ppm)	0357	144.2	0.1490	138.9	9.978	3.804	79	0.53	037.42	0
037	Zinc (ppm)	0229	144.2	6.600	138.9	9.978	3.804	79	0.54	037.41	0
037	Zinc (ppm)	0407	144.2	6.709	138.9	9.978	3.804	79	0.54	037.41	0
037	Zinc (ppm)	0038	145.5	3.000	138.9	9.978	3.804	79	0.67	037.32	0
037	Zinc (ppm)	0968	145.8	0.0200	138.9	9.978	3.804	79	0.69	037.43	0
037	Zinc (ppm)	0035	146.0	0.0000	138.9	9.978	3.804	79	0.72	037.42	0
037	Zinc (ppm)	0035	146.5	3.000	138.9	9.978	3.804	79	0.77	037.43	0
037	Zinc (ppm)	0186	146.9	7.800	138.9	9.978	3.804	79	0.81	037.52	0
037	Zinc (ppm)	0148	147.6	5.250	138.9	9.978	3.804	79	0.88	037.41	0
037	Zinc (ppm)	0265	148.0	2.000	138.9	9.978	3.804	79	0.92	037.42	0
037	Zinc (ppm)	0675	148.1	9.110	138.9	9.978	3.804	79	0.93	037.43	0
037	Zinc (ppm)	0004	148.6	2.750	138.9	9.978	3.804	79	0.98	037.41	0
037	Zinc (ppm)	0098	151.3	0.0000	138.9	9.978	3.804	79	1.25	037.41	0
037	Zinc (ppm)	0100	152.0	4.000	138.9	9.978	3.804	79	1.32	037.99	0
037	Zinc (ppm)	2188	152.4	0.0000	138.9	9.978	3.804	79	1.36	037.31	0
037	Zinc (ppm)	0098	154.0	3.100	138.9	9.978	3.804	79	1.51	037.43	0
037	Zinc (ppm)	0726	154.9	1.800	138.9	9.978	3.804	79	1.61	037.42	0
037	Zinc (ppm)	0042	155.0	14.00	138.9	9.978	3.804	79	1.62	037.42	0
037	Zinc (ppm)	0505	155.7	11.10	138.9	9.978	3.804	79	1.68	037.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
037	Zinc (ppm)	0037	166.7	1.700	138.9	9.978	3.804	79	2.79	037.42	0
037	Zinc (ppm)	0190	171.8	3.100	138.9	9.978	3.804	79	3.30	037.42	0
037	Zinc (ppm)	0354	210.2	1.750	138.9	9.978	3.804	79	7.15	037.41	0
037	Zinc (ppm)	0297	142.0	24.00	138.9	9.978	3.804	79	0.31	037.43	1
037	Zinc (ppm)	0536	25.45	0.9000	138.9	9.978	3.804	79	-11.37	037.31	2
038	Molybdenum (ppm)	0171	0.5975	0.0110	1.044	0.1913	0.0443	17	-2.33	038.41	0
038	Molybdenum (ppm)	0083	0.8000	0.0000	1.044	0.1913	0.0443	17	-1.27	038.43	0
038	Molybdenum (ppm)	0910	0.8950	0.0100	1.044	0.1913	0.0443	17	-0.78	038.52	0
038	Molybdenum (ppm)	0407	0.9092	0.0648	1.044	0.1913	0.0443	17	-0.70	038.41	0
038	Molybdenum (ppm)	0169	0.9100	0.1200	1.044	0.1913	0.0443	17	-0.70	038.43	0
038	Molybdenum (ppm)	0297	0.9500	0.1000	1.044	0.1913	0.0443	17	-0.49	038.43	0
038	Molybdenum (ppm)	0407	0.9750	0.0322	1.044	0.1913	0.0443	17	-0.36	038.43	0
038	Molybdenum (ppm)	0345	1.020	0.0200	1.044	0.1913	0.0443	17	-0.12	038.43	0
038	Molybdenum (ppm)	0186	1.020	0.0200	1.044	0.1913	0.0443	17	-0.12	038.52	0
038	Molybdenum (ppm)	0510	1.050	0.1000	1.044	0.1913	0.0443	17	0.03	038.43	0
038	Molybdenum (ppm)	0164	1.110	0.0200	1.044	0.1913	0.0443	17	0.35	038.53	0
038	Molybdenum (ppm)	0407	1.113	0.0443	1.044	0.1913	0.0443	17	0.36	038.53	0
038	Molybdenum (ppm)	0939	1.150	0.0200	1.044	0.1913	0.0443	17	0.56	038.53	0
038	Molybdenum (ppm)	0553	1.195	0.0100	1.044	0.1913	0.0443	17	0.79	038.53	0
038	Molybdenum (ppm)	0278	1.225	0.0900	1.044	0.1913	0.0443	17	0.95	038.42	0
038	Molybdenum (ppm)	0226	1.440	0.0400	1.044	0.1913	0.0443	17	2.07	038.41	0
038	Molybdenum (ppm)	0045	1.575	0.0500	1.044	0.1913	0.0443	17	2.78	038.42	0
038	Molybdenum (ppm)	0265	0.3050	0.2500	1.044	0.1913	0.0443	17	-3.86	038.42	1
040	Barium (ppm)	0407	7.748	0.1661				1		040.53	0
041	Vanadium (ppm)	0407	0.3653	0.0179				2		041.53	0
041	Vanadium (ppm)	0553	0.3665	0.0050				2		041.53	0
042	Chloride (%)	0629	0.4750	0.0100	0.4980	0.0267	0.0650	6	-0.86	042.00	0
042	Chloride (%)	0948	0.4800	0.0000	0.4980	0.0267	0.0650	6	-0.67	042.00	0
042	Chloride (%)	0226	0.4850	0.0100	0.4980	0.0267	0.0650	6	-0.49	042.00	0
042	Chloride (%)	0297	0.5000	0.0600	0.4980	0.0267	0.0650	6	0.07	042.99	0
042	Chloride (%)	2188	0.5100	0.0000	0.4980	0.0267	0.0650	6	0.45	042.01	0
042	Chloride (%)	2302	0.5550	0.3100	0.4980	0.0267	0.0650	6	2.13	042.99	0
042	Chloride (%)	0941	4.380	0.3600	0.4980	0.0267	0.0650	6	145.14	042.02	2
099	Menadione (form) (ppm)	2319	< 0.5					0		099.01	5
101	Choline Chloride (ppm)	0227	2,350	0.0000				1		101.99	0
102	Niacin (ppm)	2319	65.70	0.4000				2		102.02	0
102	Niacin (ppm)	0227	101.0	4.000				2		102.01	0
103	Pantothenic Acid (ppm)	0227	20.80	0.4000				1		103.01	0
104	Riboflavin (ppm)	0910	7.455	0.5100				3		104.03	0
104	Riboflavin (ppm)	2319	7.740	0.3800				3		104.03	0
104	Riboflavin (ppm)	0227	11.70	1.200				3		104.00	0
105	Thiamine (ppm)	0910	4.000	3.000				2		105.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
105	Thiamine (ppm)	0227	8.535	0.7100				2		105.01	0
106	Vitamin A (KU / kg)	0038	6.715	1.650	8.913	1.288	1.161	8	-1.71	106.02	0
106	Vitamin A (KU / kg)	0619	7.890	0.5000	8.913	1.288	1.161	8	-0.79	106.02	0
106	Vitamin A (KU / kg)	0563	8.724	0.9054	8.913	1.288	1.161	8	-0.15	106.02	0
106	Vitamin A (KU / kg)	0227	8.895	0.5900	8.913	1.288	1.161	8	-0.01	106.00	0
106	Vitamin A (KU / kg)	2319	9.155	0.7300	8.913	1.288	1.161	8	0.19	106.02	0
106	Vitamin A (KU / kg)	0148	9.215	2.010	8.913	1.288	1.161	8	0.23	106.02	0
106	Vitamin A (KU / kg)	0689	9.600	1.000	8.913	1.288	1.161	8	0.53	106.02	0
106	Vitamin A (KU / kg)	0098	13.15	1.900	8.913	1.288	1.161	8	3.29	106.01	0
106	Vitamin A (KU / kg)	2302	11,325	1,690	8.913	1.288	1.161	8	8785.49	106.01	2
107	Vitamin B12 (ppb)	0227	47.35	4.500				1		107.00	0
108	Vitamin D3 (KU / kg)	0619	2.660	0.6000	4.405	2.438	0.6050	4		108.02	0
108	Vitamin D3 (KU / kg)	0010	3.165	1.130	4.405	2.438	0.6050	4		108.99	0
108	Vitamin D3 (KU / kg)	0227	3.800	0.0600	4.405	2.438	0.6050	4		108.02	0
108	Vitamin D3 (KU / kg)	2319	7.995	0.6300	4.405	2.438	0.6050	4		108.02	0
109	Vitamin E (IU / kg)	0619	12.50	1.000	39.79	14.48	2.362	8	-1.88	109.02	0
109	Vitamin E (IU / kg)	2302	20.38	3.500	39.79	14.48	2.362	8	-1.34	109.99	0
109	Vitamin E (IU / kg)	0148	44.55	2.500	39.79	14.48	2.362	8	0.33	109.02	0
109	Vitamin E (IU / kg)	0098	45.30	2.800	39.79	14.48	2.362	8	0.38	109.02	0
109	Vitamin E (IU / kg)	2193	46.51	6.300	39.79	14.48	2.362	8	0.46	109.02	0
109	Vitamin E (IU / kg)	2319	47.25	0.3300	39.79	14.48	2.362	8	0.52	109.02	0
109	Vitamin E (IU / kg)	0227	48.00	2.200	39.79	14.48	2.362	8	0.57	109.02	0
109	Vitamin E (IU / kg)	0563	48.24	0.2683	39.79	14.48	2.362	8	0.58	109.02	0
111	Vitamin C, Phosphorylated (ppm)	0227	< 4.4					0		111.00	5
112	Pyridoxine (µg / g)	2319	6.910	0.0400				2		112.01	0
112	Pyridoxine (µg / g)	0227	8.415	0.4500				2		112.01	0
113	Folic Acid (ppm)	2319	0.9100	0.0200				2		113.02	0
113	Folic Acid (ppm)	0227	1.410	0.1600				2		113.01	0
114	Biotin (ppm)	0227	0.4925	0.0290				1		114.01	0
115	Non Protein N (NPN) (%)	2129	0.4558	0.0828				2		115.99	0
115	Non Protein N (NPN) (%)	0098	0.6300	0.0000				2		115.00	0
118	Peroxide value (meq/kg)	0853	5.800	1.400				1		118.99	0
120	Alanine (%)	0353	1.000	0.0200	1.098	0.0391	0.0115	28	-2.51	120.00	0
120	Alanine (%)	2319	1.000	0.0000	1.098	0.0391	0.0115	28	-2.51	120.05	0
120	Alanine (%)	0968	1.015	0.0100	1.098	0.0391	0.0115	28	-2.13	120.00	0
120	Alanine (%)	0148	1.040	0.0200	1.098	0.0391	0.0115	28	-1.49	120.05	0
120	Alanine (%)	2246	1.049	0.0008	1.098	0.0391	0.0115	28	-1.27	120.05	0
120	Alanine (%)	0939	1.070	0.0200	1.098	0.0391	0.0115	28	-0.72	120.00	0
120	Alanine (%)	0918	1.080	0.0200	1.098	0.0391	0.0115	28	-0.46	120.00	0
120	Alanine (%)	2196	1.080	0.0000	1.098	0.0391	0.0115	28	-0.46	120.05	0
120	Alanine (%)	0941	1.086	0.0000	1.098	0.0391	0.0115	28	-0.31	120.00	0
120	Alanine (%)	0407	1.087	0.0020	1.098	0.0391	0.0115	28	-0.29	120.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
120	Alanine (%)	0652	1.090	0.0000	1.098	0.0391	0.0115	28	-0.21	120.00	0
120	Alanine (%)	2059	1.094	0.0080	1.098	0.0391	0.0115	28	-0.11	120.00	0
120	Alanine (%)	2146	1.095	0.0300	1.098	0.0391	0.0115	28	-0.08	120.99	0
120	Alanine (%)	0684	1.100	0.0000	1.098	0.0391	0.0115	28	0.05	120.00	0
120	Alanine (%)	2339	1.104	0.0140	1.098	0.0391	0.0115	28	0.15	120.00	0
120	Alanine (%)	2193	1.110	0.0100	1.098	0.0391	0.0115	28	0.30	120.05	0
120	Alanine (%)	0504	1.110	0.0200	1.098	0.0391	0.0115	28	0.30	120.00	0
120	Alanine (%)	0571	1.114	0.0000	1.098	0.0391	0.0115	28	0.40	120.00	0
120	Alanine (%)	0227	1.115	0.0100	1.098	0.0391	0.0115	28	0.43	120.99	0
120	Alanine (%)	0226	1.115	0.0208	1.098	0.0391	0.0115	28	0.44	120.00	0
120	Alanine (%)	0644	1.124	0.0000	1.098	0.0391	0.0115	28	0.66	120.00	0
120	Alanine (%)	0619	1.125	0.0100	1.098	0.0391	0.0115	28	0.69	120.00	0
120	Alanine (%)	2188	1.127	0.0000	1.098	0.0391	0.0115	28	0.74	120.05	0
120	Alanine (%)	0870	1.131	0.0156	1.098	0.0391	0.0115	28	0.85	120.00	0
120	Alanine (%)	0354	1.134	0.0090	1.098	0.0391	0.0115	28	0.90	120.00	0
120	Alanine (%)	0859	1.137	0.0257	1.098	0.0391	0.0115	28	1.00	120.00	0
120	Alanine (%)	0675	1.185	0.0100	1.098	0.0391	0.0115	28	2.22	120.00	0
120	Alanine (%)	0098	1.193	0.0450	1.098	0.0391	0.0115	28	2.41	120.02	0
120	Alanine (%)	0910	0.9700	0.2400	1.098	0.0391	0.0115	28	-3.28	120.00	1
121	Arginine (%)	0407	1.114	0.0020	1.235	0.0472	0.0238	27	-2.56	121.05	0
121	Arginine (%)	0148	1.165	0.0100	1.235	0.0472	0.0238	27	-1.48	121.05	0
121	Arginine (%)	0353	1.180	0.0200	1.235	0.0472	0.0238	27	-1.16	121.00	0
121	Arginine (%)	2319	1.190	0.0400	1.235	0.0472	0.0238	27	-0.95	121.05	0
121	Arginine (%)	0939	1.195	0.0300	1.235	0.0472	0.0238	27	-0.84	121.00	0
121	Arginine (%)	0968	1.195	0.0300	1.235	0.0472	0.0238	27	-0.84	121.00	0
121	Arginine (%)	0354	1.199	0.0070	1.235	0.0472	0.0238	27	-0.77	121.00	0
121	Arginine (%)	0859	1.207	0.0335	1.235	0.0472	0.0238	27	-0.60	121.00	0
121	Arginine (%)	0098	1.212	0.0110	1.235	0.0472	0.0238	27	-0.49	121.02	0
121	Arginine (%)	0684	1.220	0.0000	1.235	0.0472	0.0238	27	-0.31	121.00	0
121	Arginine (%)	0227	1.220	0.0400	1.235	0.0472	0.0238	27	-0.31	121.99	0
121	Arginine (%)	2193	1.225	0.0440	1.235	0.0472	0.0238	27	-0.21	121.05	0
121	Arginine (%)	2339	1.231	0.0300	1.235	0.0472	0.0238	27	-0.08	121.00	0
121	Arginine (%)	0918	1.235	0.0300	1.235	0.0472	0.0238	27	0.01	121.00	0
121	Arginine (%)	2196	1.237	0.0000	1.235	0.0472	0.0238	27	0.05	121.05	0
121	Arginine (%)	0504	1.240	0.0200	1.235	0.0472	0.0238	27	0.11	121.00	0
121	Arginine (%)	0870	1.242	0.0631	1.235	0.0472	0.0238	27	0.15	121.00	0
121	Arginine (%)	0941	1.244	0.0000	1.235	0.0472	0.0238	27	0.20	121.00	0
121	Arginine (%)	0652	1.250	0.0200	1.235	0.0472	0.0238	27	0.32	121.00	0
121	Arginine (%)	0226	1.258	0.0153	1.235	0.0472	0.0238	27	0.50	121.00	0
121	Arginine (%)	2246	1.266	0.0185	1.235	0.0472	0.0238	27	0.67	121.05	0
121	Arginine (%)	0644	1.276	0.0000	1.235	0.0472	0.0238	27	0.88	121.00	0
121	Arginine (%)	2059	1.280	0.0060	1.235	0.0472	0.0238	27	0.96	121.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
121	Arginine (%)	0571	1.297	0.0330	1.235	0.0472	0.0238	27	1.31	121.00	0
121	Arginine (%)	0619	1.300	0.1000	1.235	0.0472	0.0238	27	1.38	121.00	0
121	Arginine (%)	0675	1.320	0.0400	1.235	0.0472	0.0238	27	1.81	121.00	0
121	Arginine (%)	2188	1.351	0.0000	1.235	0.0472	0.0238	27	2.46	121.05	0
121	Arginine (%)	0910	1.105	0.2500	1.235	0.0472	0.0238	27	-2.75	121.00	1
122	Aspartic (%)	2196	1.564	0.0000	1.775	0.0653	0.0257	28	-3.23	122.05	0
122	Aspartic (%)	2246	1.571	0.0611	1.775	0.0653	0.0257	28	-3.12	122.05	0
122	Aspartic (%)	2319	1.670	0.0200	1.775	0.0653	0.0257	28	-1.60	122.05	0
122	Aspartic (%)	0941	1.686	0.0000	1.775	0.0653	0.0257	28	-1.36	122.00	0
122	Aspartic (%)	0918	1.715	0.0700	1.775	0.0653	0.0257	28	-0.91	122.00	0
122	Aspartic (%)	0939	1.735	0.0100	1.775	0.0653	0.0257	28	-0.61	122.00	0
122	Aspartic (%)	0148	1.735	0.0100	1.775	0.0653	0.0257	28	-0.61	122.05	0
122	Aspartic (%)	0968	1.740	0.0200	1.775	0.0653	0.0257	28	-0.53	122.00	0
122	Aspartic (%)	0353	1.750	0.0000	1.775	0.0653	0.0257	28	-0.38	122.00	0
122	Aspartic (%)	0684	1.750	0.0200	1.775	0.0653	0.0257	28	-0.38	122.00	0
122	Aspartic (%)	0354	1.757	0.0840	1.775	0.0653	0.0257	28	-0.27	122.00	0
122	Aspartic (%)	2146	1.760	0.0200	1.775	0.0653	0.0257	28	-0.22	122.99	0
122	Aspartic (%)	2339	1.774	0.0140	1.775	0.0653	0.0257	28	-0.01	122.00	0
122	Aspartic (%)	0619	1.775	0.0100	1.775	0.0653	0.0257	28	0.01	122.00	0
122	Aspartic (%)	0227	1.780	0.0400	1.775	0.0653	0.0257	28	0.08	122.99	0
122	Aspartic (%)	2059	1.784	0.0180	1.775	0.0653	0.0257	28	0.14	122.00	0
122	Aspartic (%)	0226	1.787	0.0275	1.775	0.0653	0.0257	28	0.19	122.00	0
122	Aspartic (%)	0504	1.795	0.0100	1.775	0.0653	0.0257	28	0.31	122.00	0
122	Aspartic (%)	2188	1.795	0.0000	1.775	0.0653	0.0257	28	0.31	122.05	0
122	Aspartic (%)	0644	1.803	0.0000	1.775	0.0653	0.0257	28	0.43	122.00	0
122	Aspartic (%)	0652	1.810	0.0600	1.775	0.0653	0.0257	28	0.54	122.00	0
122	Aspartic (%)	0571	1.814	0.0150	1.775	0.0653	0.0257	28	0.59	122.00	0
122	Aspartic (%)	0859	1.818	0.0517	1.775	0.0653	0.0257	28	0.66	122.00	0
122	Aspartic (%)	2193	1.850	0.0410	1.775	0.0653	0.0257	28	1.15	122.05	0
122	Aspartic (%)	0098	1.851	0.0310	1.775	0.0653	0.0257	28	1.16	122.02	0
122	Aspartic (%)	0870	1.853	0.0497	1.775	0.0653	0.0257	28	1.21	122.00	0
122	Aspartic (%)	0407	1.890	0.0060	1.775	0.0653	0.0257	28	1.77	122.05	0
122	Aspartic (%)	0675	1.955	0.0300	1.775	0.0653	0.0257	28	2.76	122.00	0
122	Aspartic (%)	0910	1.585	0.3300	1.775	0.0653	0.0257	28	-2.90	122.00	1
124	Cysteine/Cystine (%)	2246	0.2442	0.0405	0.2858	0.0264	0.0103	28	-1.57	124.05	0
124	Cysteine/Cystine (%)	2188	0.2500	0.0000	0.2858	0.0264	0.0103	28	-1.35	124.05	0
124	Cysteine/Cystine (%)	2319	0.2550	0.0100	0.2858	0.0264	0.0103	28	-1.16	124.05	0
124	Cysteine/Cystine (%)	0619	0.2595	0.0030	0.2858	0.0264	0.0103	28	-0.99	124.00	0
124	Cysteine/Cystine (%)	0918	0.2600	0.0400	0.2858	0.0264	0.0103	28	-0.97	124.00	0
124	Cysteine/Cystine (%)	0227	0.2600	0.0000	0.2858	0.0264	0.0103	28	-0.97	124.99	0
124	Cysteine/Cystine (%)	0870	0.2609	0.0094	0.2858	0.0264	0.0103	28	-0.94	124.00	0
124	Cysteine/Cystine (%)	2059	0.2685	0.0010	0.2858	0.0264	0.0103	28	-0.65	124.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
124	Cysteine/Cystine (%)	0226	0.2717	0.0135	0.2858	0.0264	0.0103	28	-0.53	124.00	0
124	Cysteine/Cystine (%)	0353	0.2750	0.0100	0.2858	0.0264	0.0103	28	-0.41	124.00	0
124	Cysteine/Cystine (%)	0644	0.2810	0.0000	0.2858	0.0264	0.0103	28	-0.18	124.00	0
124	Cysteine/Cystine (%)	0859	0.2811	0.0069	0.2858	0.0264	0.0103	28	-0.18	124.00	0
124	Cysteine/Cystine (%)	0652	0.2850	0.0100	0.2858	0.0264	0.0103	28	-0.03	124.00	0
124	Cysteine/Cystine (%)	0939	0.2850	0.0100	0.2858	0.0264	0.0103	28	-0.03	124.00	0
124	Cysteine/Cystine (%)	0407	0.2855	0.0130	0.2858	0.0264	0.0103	28	-0.01	124.05	0
124	Cysteine/Cystine (%)	0148	0.2915	0.0330	0.2858	0.0264	0.0103	28	0.22	124.05	0
124	Cysteine/Cystine (%)	0941	0.2940	0.0000	0.2858	0.0264	0.0103	28	0.31	124.00	0
124	Cysteine/Cystine (%)	0684	0.2950	0.0100	0.2858	0.0264	0.0103	28	0.35	124.00	0
124	Cysteine/Cystine (%)	0968	0.2950	0.0100	0.2858	0.0264	0.0103	28	0.35	124.00	0
124	Cysteine/Cystine (%)	0098	0.2955	0.0030	0.2858	0.0264	0.0103	28	0.37	124.02	0
124	Cysteine/Cystine (%)	2193	0.2965	0.0110	0.2858	0.0264	0.0103	28	0.41	124.05	0
124	Cysteine/Cystine (%)	0571	0.3015	0.0070	0.2858	0.0264	0.0103	28	0.60	124.00	0
124	Cysteine/Cystine (%)	2196	0.3020	0.0000	0.2858	0.0264	0.0103	28	0.61	124.05	0
124	Cysteine/Cystine (%)	2146	0.3050	0.0100	0.2858	0.0264	0.0103	28	0.73	124.99	0
124	Cysteine/Cystine (%)	0504	0.3250	0.0100	0.2858	0.0264	0.0103	28	1.48	124.00	0
124	Cysteine/Cystine (%)	0675	0.3250	0.0100	0.2858	0.0264	0.0103	28	1.48	124.00	0
124	Cysteine/Cystine (%)	2339	0.3285	0.0110	0.2858	0.0264	0.0103	28	1.62	124.00	0
124	Cysteine/Cystine (%)	0354	0.3305	0.0050	0.2858	0.0264	0.0103	28	1.69	124.00	0
124	Cysteine/Cystine (%)	0910	0.2550	0.0700	0.2858	0.0264	0.0103	28	-1.16	124.00	1
125	Glutamic (%)	2196	3.022	0.0000	3.262	0.1311	0.0364	28	-1.83	125.05	0
125	Glutamic (%)	2319	3.045	0.0100	3.262	0.1311	0.0364	28	-1.65	125.05	0
125	Glutamic (%)	2146	3.060	0.0000	3.262	0.1311	0.0364	28	-1.54	125.99	0
125	Glutamic (%)	0918	3.080	0.0000	3.262	0.1311	0.0364	28	-1.39	125.00	0
125	Glutamic (%)	0968	3.090	0.0600	3.262	0.1311	0.0364	28	-1.31	125.00	0
125	Glutamic (%)	2246	3.116	0.0531	3.262	0.1311	0.0364	28	-1.11	125.05	0
125	Glutamic (%)	0148	3.205	0.0100	3.262	0.1311	0.0364	28	-0.43	125.05	0
125	Glutamic (%)	0941	3.216	0.0000	3.262	0.1311	0.0364	28	-0.35	125.00	0
125	Glutamic (%)	2193	3.236	0.0970	3.262	0.1311	0.0364	28	-0.20	125.05	0
125	Glutamic (%)	0939	3.250	0.0000	3.262	0.1311	0.0364	28	-0.09	125.00	0
125	Glutamic (%)	2059	3.250	0.0100	3.262	0.1311	0.0364	28	-0.09	125.00	0
125	Glutamic (%)	0619	3.265	0.0500	3.262	0.1311	0.0364	28	0.02	125.00	0
125	Glutamic (%)	0098	3.269	0.0490	3.262	0.1311	0.0364	28	0.05	125.02	0
125	Glutamic (%)	0652	3.285	0.0900	3.262	0.1311	0.0364	28	0.18	125.00	0
125	Glutamic (%)	0859	3.299	0.0922	3.262	0.1311	0.0364	28	0.28	125.00	0
125	Glutamic (%)	0684	3.300	0.0200	3.262	0.1311	0.0364	28	0.29	125.00	0
125	Glutamic (%)	0227	3.300	0.0400	3.262	0.1311	0.0364	28	0.29	125.99	0
125	Glutamic (%)	2339	3.304	0.0010	3.262	0.1311	0.0364	28	0.32	125.00	0
125	Glutamic (%)	0353	3.305	0.0100	3.262	0.1311	0.0364	28	0.33	125.00	0
125	Glutamic (%)	0354	3.309	0.1720	3.262	0.1311	0.0364	28	0.36	125.00	0
125	Glutamic (%)	0571	3.314	0.0090	3.262	0.1311	0.0364	28	0.39	125.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
125	Glutamic (%)	0504	3.335	0.0300	3.262	0.1311	0.0364	28	0.56	125.00	0
125	Glutamic (%)	0644	3.336	0.0000	3.262	0.1311	0.0364	28	0.57	125.00	0
125	Glutamic (%)	0870	3.344	0.1028	3.262	0.1311	0.0364	28	0.63	125.00	0
125	Glutamic (%)	0407	3.363	0.0090	3.262	0.1311	0.0364	28	0.77	125.05	0
125	Glutamic (%)	2188	3.452	0.0000	3.262	0.1311	0.0364	28	1.45	125.05	0
125	Glutamic (%)	0226	3.465	0.0631	3.262	0.1311	0.0364	28	1.55	125.00	0
125	Glutamic (%)	0675	3.490	0.0400	3.262	0.1311	0.0364	28	1.74	125.00	0
125	Glutamic (%)	0910	2.875	0.6100	3.262	0.1311	0.0364	28	-2.95	125.00	1
126	Glycine (%)	2196	1.017	0.0000	1.138	0.0549	0.0146	28	-2.20	126.05	0
126	Glycine (%)	2319	1.040	0.0000	1.138	0.0549	0.0146	28	-1.78	126.05	0
126	Glycine (%)	0407	1.045	0.0020	1.138	0.0549	0.0146	28	-1.69	126.05	0
126	Glycine (%)	0148	1.050	0.0200	1.138	0.0549	0.0146	28	-1.60	126.05	0
126	Glycine (%)	0918	1.075	0.0100	1.138	0.0549	0.0146	28	-1.14	126.00	0
126	Glycine (%)	2193	1.098	0.0100	1.138	0.0549	0.0146	28	-0.72	126.05	0
126	Glycine (%)	0968	1.100	0.0200	1.138	0.0549	0.0146	28	-0.69	126.00	0
126	Glycine (%)	0353	1.105	0.0100	1.138	0.0549	0.0146	28	-0.60	126.00	0
126	Glycine (%)	0652	1.110	0.0000	1.138	0.0549	0.0146	28	-0.50	126.00	0
126	Glycine (%)	2339	1.126	0.0100	1.138	0.0549	0.0146	28	-0.21	126.00	0
126	Glycine (%)	0941	1.129	0.0000	1.138	0.0549	0.0146	28	-0.16	126.00	0
126	Glycine (%)	0684	1.135	0.0100	1.138	0.0549	0.0146	28	-0.05	126.00	0
126	Glycine (%)	0939	1.140	0.0200	1.138	0.0549	0.0146	28	0.04	126.00	0
126	Glycine (%)	2059	1.146	0.0060	1.138	0.0549	0.0146	28	0.15	126.00	0
126	Glycine (%)	0504	1.150	0.0200	1.138	0.0549	0.0146	28	0.22	126.00	0
126	Glycine (%)	0354	1.153	0.0400	1.138	0.0549	0.0146	28	0.28	126.00	0
126	Glycine (%)	0227	1.155	0.0100	1.138	0.0549	0.0146	28	0.32	126.99	0
126	Glycine (%)	2246	1.159	0.0276	1.138	0.0549	0.0146	28	0.38	126.05	0
126	Glycine (%)	0870	1.159	0.0591	1.138	0.0549	0.0146	28	0.39	126.00	0
126	Glycine (%)	0859	1.162	0.0175	1.138	0.0549	0.0146	28	0.45	126.00	0
126	Glycine (%)	0571	1.172	0.0100	1.138	0.0549	0.0146	28	0.62	126.00	0
126	Glycine (%)	2146	1.180	0.0400	1.138	0.0549	0.0146	28	0.77	126.99	0
126	Glycine (%)	0226	1.184	0.0114	1.138	0.0549	0.0146	28	0.84	126.00	0
126	Glycine (%)	0619	1.185	0.0100	1.138	0.0549	0.0146	28	0.86	126.00	0
126	Glycine (%)	0644	1.190	0.0000	1.138	0.0549	0.0146	28	0.95	126.00	0
126	Glycine (%)	0098	1.191	0.0260	1.138	0.0549	0.0146	28	0.97	126.02	0
126	Glycine (%)	0675	1.210	0.0200	1.138	0.0549	0.0146	28	1.32	126.00	0
126	Glycine (%)	2188	1.220	0.0000	1.138	0.0549	0.0146	28	1.50	126.05	0
126	Glycine (%)	0910	1.015	0.2300	1.138	0.0549	0.0146	28	-2.23	126.00	1
127	Histidine (%)	0407	0.4180	0.0020	0.4952	0.0259	0.0108	28	-2.97	127.05	0
127	Histidine (%)	0354	0.4630	0.0120	0.4952	0.0259	0.0108	28	-1.24	127.00	0
127	Histidine (%)	2193	0.4635	0.0050	0.4952	0.0259	0.0108	28	-1.22	127.05	0
127	Histidine (%)	0918	0.4650	0.0100	0.4952	0.0259	0.0108	28	-1.16	127.00	0
127	Histidine (%)	0226	0.4699	0.0172	0.4952	0.0259	0.0108	28	-0.97	127.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
127	Histidine (%)	0353	0.4750	0.0300	0.4952	0.0259	0.0108	28	-0.78	127.00	0
127	Histidine (%)	0968	0.4750	0.0100	0.4952	0.0259	0.0108	28	-0.78	127.00	0
127	Histidine (%)	0859	0.4761	0.0133	0.4952	0.0259	0.0108	28	-0.74	127.00	0
127	Histidine (%)	0939	0.4800	0.0000	0.4952	0.0259	0.0108	28	-0.58	127.00	0
127	Histidine (%)	0148	0.4805	0.0050	0.4952	0.0259	0.0108	28	-0.57	127.05	0
127	Histidine (%)	0098	0.4845	0.0050	0.4952	0.0259	0.0108	28	-0.41	127.02	0
127	Histidine (%)	0684	0.4850	0.0100	0.4952	0.0259	0.0108	28	-0.39	127.00	0
127	Histidine (%)	2319	0.4900	0.0200	0.4952	0.0259	0.0108	28	-0.20	127.05	0
127	Histidine (%)	2339	0.4960	0.0180	0.4952	0.0259	0.0108	28	0.03	127.00	0
127	Histidine (%)	2059	0.4975	0.0010	0.4952	0.0259	0.0108	28	0.09	127.00	0
127	Histidine (%)	2196	0.4990	0.0000	0.4952	0.0259	0.0108	28	0.15	127.05	0
127	Histidine (%)	0571	0.5000	0.0080	0.4952	0.0259	0.0108	28	0.19	127.00	0
127	Histidine (%)	0644	0.5080	0.0000	0.4952	0.0259	0.0108	28	0.49	127.00	0
127	Histidine (%)	0675	0.5100	0.0200	0.4952	0.0259	0.0108	28	0.57	127.00	0
127	Histidine (%)	0227	0.5100	0.0000	0.4952	0.0259	0.0108	28	0.57	127.99	0
127	Histidine (%)	0870	0.5120	0.0391	0.4952	0.0259	0.0108	28	0.65	127.00	0
127	Histidine (%)	2188	0.5140	0.0000	0.4952	0.0259	0.0108	28	0.73	127.05	0
127	Histidine (%)	0504	0.5150	0.0100	0.4952	0.0259	0.0108	28	0.76	127.00	0
127	Histidine (%)	2146	0.5200	0.0200	0.4952	0.0259	0.0108	28	0.96	127.99	0
127	Histidine (%)	0652	0.5250	0.0100	0.4952	0.0259	0.0108	28	1.15	127.00	0
127	Histidine (%)	2246	0.5295	0.0280	0.4952	0.0259	0.0108	28	1.32	127.05	0
127	Histidine (%)	0941	0.5310	0.0000	0.4952	0.0259	0.0108	28	1.38	127.00	0
127	Histidine (%)	0619	0.5990	0.0080	0.4952	0.0259	0.0108	28	4.00	127.00	0
127	Histidine (%)	0910	0.4350	0.1300	0.4952	0.0259	0.0108	28	-2.32	127.00	1
128	Isoleucine (%)	2319	0.5450	0.0300	0.7399	0.0569	0.0198	28	-3.43	128.05	0
128	Isoleucine (%)	0353	0.6550	0.0300	0.7399	0.0569	0.0198	28	-1.49	128.00	0
128	Isoleucine (%)	0870	0.6780	0.0371	0.7399	0.0569	0.0198	28	-1.09	128.00	0
128	Isoleucine (%)	0354	0.6825	0.0230	0.7399	0.0569	0.0198	28	-1.01	128.00	0
128	Isoleucine (%)	0918	0.6850	0.0100	0.7399	0.0569	0.0198	28	-0.96	128.00	0
128	Isoleucine (%)	0939	0.6850	0.0100	0.7399	0.0569	0.0198	28	-0.96	128.00	0
128	Isoleucine (%)	0968	0.6900	0.0200	0.7399	0.0569	0.0198	28	-0.88	128.00	0
128	Isoleucine (%)	2196	0.7200	0.0000	0.7399	0.0569	0.0198	28	-0.35	128.05	0
128	Isoleucine (%)	0148	0.7230	0.0120	0.7399	0.0569	0.0198	28	-0.30	128.05	0
128	Isoleucine (%)	0684	0.7250	0.0100	0.7399	0.0569	0.0198	28	-0.26	128.00	0
128	Isoleucine (%)	0407	0.7295	0.0010	0.7399	0.0569	0.0198	28	-0.18	128.05	0
128	Isoleucine (%)	0652	0.7300	0.0200	0.7399	0.0569	0.0198	28	-0.17	128.00	0
128	Isoleucine (%)	0941	0.7300	0.0000	0.7399	0.0569	0.0198	28	-0.17	128.00	0
128	Isoleucine (%)	0619	0.7315	0.0350	0.7399	0.0569	0.0198	28	-0.15	128.00	0
128	Isoleucine (%)	0226	0.7339	0.0212	0.7399	0.0569	0.0198	28	-0.11	128.00	0
128	Isoleucine (%)	2146	0.7350	0.0300	0.7399	0.0569	0.0198	28	-0.09	128.99	0
128	Isoleucine (%)	0098	0.7425	0.0710	0.7399	0.0569	0.0198	28	0.05	128.02	0
128	Isoleucine (%)	0227	0.7500	0.0200	0.7399	0.0569	0.0198	28	0.18	128.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
128	Isoleucine (%)	0859	0.7531	0.0129	0.7399	0.0569	0.0198	28	0.23	128.00	0
128	Isoleucine (%)	0571	0.7540	0.0160	0.7399	0.0569	0.0198	28	0.25	128.00	0
128	Isoleucine (%)	2059	0.7705	0.0070	0.7399	0.0569	0.0198	28	0.54	128.00	0
128	Isoleucine (%)	2339	0.7720	0.0040	0.7399	0.0569	0.0198	28	0.56	128.00	0
128	Isoleucine (%)	0644	0.8000	0.0000	0.7399	0.0569	0.0198	28	1.06	128.00	0
128	Isoleucine (%)	2188	0.8030	0.0000	0.7399	0.0569	0.0198	28	1.11	128.05	0
128	Isoleucine (%)	0675	0.8100	0.0200	0.7399	0.0569	0.0198	28	1.23	128.00	0
128	Isoleucine (%)	2246	0.8233	0.0464	0.7399	0.0569	0.0198	28	1.47	128.05	0
128	Isoleucine (%)	0504	0.8250	0.0300	0.7399	0.0569	0.0198	28	1.50	128.00	0
128	Isoleucine (%)	2193	0.8255	0.0390	0.7399	0.0569	0.0198	28	1.51	128.05	0
128	Isoleucine (%)	0910	0.6750	0.1500	0.7399	0.0569	0.0198	28	-1.14	128.00	1
129	Leucine (%)	2196	1.326	0.0000	1.562	0.0583	0.0208	28	-4.04	129.05	0
129	Leucine (%)	2319	1.390	0.0000	1.562	0.0583	0.0208	28	-2.95	129.05	0
129	Leucine (%)	0407	1.485	0.0020	1.562	0.0583	0.0208	28	-1.32	129.05	0
129	Leucine (%)	0968	1.490	0.0200	1.562	0.0583	0.0208	28	-1.23	129.00	0
129	Leucine (%)	0939	1.500	0.0200	1.562	0.0583	0.0208	28	-1.06	129.00	0
129	Leucine (%)	0918	1.510	0.0000	1.562	0.0583	0.0208	28	-0.89	129.00	0
129	Leucine (%)	0148	1.510	0.0600	1.562	0.0583	0.0208	28	-0.89	129.05	0
129	Leucine (%)	2246	1.519	0.0234	1.562	0.0583	0.0208	28	-0.73	129.05	0
129	Leucine (%)	0098	1.544	0.0790	1.562	0.0583	0.0208	28	-0.31	129.02	0
129	Leucine (%)	0941	1.548	0.0000	1.562	0.0583	0.0208	28	-0.23	129.00	0
129	Leucine (%)	0684	1.550	0.0200	1.562	0.0583	0.0208	28	-0.20	129.00	0
129	Leucine (%)	2193	1.556	0.0350	1.562	0.0583	0.0208	28	-0.11	129.05	0
129	Leucine (%)	0870	1.558	0.0240	1.562	0.0583	0.0208	28	-0.07	129.00	0
129	Leucine (%)	0652	1.560	0.0000	1.562	0.0583	0.0208	28	-0.03	129.00	0
129	Leucine (%)	2339	1.567	0.0140	1.562	0.0583	0.0208	28	0.09	129.00	0
129	Leucine (%)	0354	1.574	0.0580	1.562	0.0583	0.0208	28	0.21	129.00	0
129	Leucine (%)	0571	1.576	0.0160	1.562	0.0583	0.0208	28	0.25	129.00	0
129	Leucine (%)	2059	1.581	0.0050	1.562	0.0583	0.0208	28	0.32	129.00	0
129	Leucine (%)	2146	1.585	0.0500	1.562	0.0583	0.0208	28	0.40	129.99	0
129	Leucine (%)	0227	1.600	0.0400	1.562	0.0583	0.0208	28	0.66	129.99	0
129	Leucine (%)	0859	1.601	0.0469	1.562	0.0583	0.0208	28	0.67	129.00	0
129	Leucine (%)	0353	1.605	0.0100	1.562	0.0583	0.0208	28	0.74	129.00	0
129	Leucine (%)	0226	1.609	0.0078	1.562	0.0583	0.0208	28	0.82	129.00	0
129	Leucine (%)	0644	1.614	0.0000	1.562	0.0583	0.0208	28	0.90	129.00	0
129	Leucine (%)	0504	1.615	0.0300	1.562	0.0583	0.0208	28	0.91	129.00	0
129	Leucine (%)	0619	1.625	0.0100	1.562	0.0583	0.0208	28	1.09	129.00	0
129	Leucine (%)	0675	1.685	0.0100	1.562	0.0583	0.0208	28	2.12	129.00	0
129	Leucine (%)	2188	1.687	0.0000	1.562	0.0583	0.0208	28	2.15	129.05	0
129	Leucine (%)	0910	1.370	0.2800	1.562	0.0583	0.0208	28	-3.29	129.00	1
130	L-Lysine (%)	2188	0.7300	0.0000	0.9917	0.0592	0.0208	32	-4.42	130.05	0
130	L-Lysine (%)	0848	0.8650	0.0100	0.9917	0.0592	0.0208	32	-2.14	130.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
130	L-Lysine (%)	2319	0.8700	0.0800	0.9917	0.0592	0.0208	32	-2.05	130.05	0
130	L-Lysine (%)	0148	0.9000	0.0020	0.9917	0.0592	0.0208	32	-1.55	130.05	0
130	L-Lysine (%)	2246	0.9252	0.0059	0.9917	0.0592	0.0208	32	-1.12	130.05	0
130	L-Lysine (%)	0918	0.9350	0.0100	0.9917	0.0592	0.0208	32	-0.96	130.00	0
130	L-Lysine (%)	2322	0.9490	0.0040	0.9917	0.0592	0.0208	32	-0.72	130.00	0
130	L-Lysine (%)	0939	0.9550	0.0100	0.9917	0.0592	0.0208	32	-0.62	130.00	0
130	L-Lysine (%)	2144	0.9600	0.0000	0.9917	0.0592	0.0208	32	-0.53	130.99	0
130	L-Lysine (%)	0512	0.9614	0.0283	0.9917	0.0592	0.0208	32	-0.51	130.00	0
130	L-Lysine (%)	0353	0.9650	0.0100	0.9917	0.0592	0.0208	32	-0.45	130.00	0
130	L-Lysine (%)	0968	0.9700	0.0200	0.9917	0.0592	0.0208	32	-0.37	130.00	0
130	L-Lysine (%)	0941	0.9790	0.0000	0.9917	0.0592	0.0208	32	-0.21	130.00	0
130	L-Lysine (%)	0684	0.9850	0.0100	0.9917	0.0592	0.0208	32	-0.11	130.00	0
130	L-Lysine (%)	0354	0.9910	0.0260	0.9917	0.0592	0.0208	32	-0.01	130.00	0
130	L-Lysine (%)	0652	1.000	0.0200	0.9917	0.0592	0.0208	32	0.14	130.00	0
130	L-Lysine (%)	2339	1.004	0.0060	0.9917	0.0592	0.0208	32	0.21	130.00	0
130	L-Lysine (%)	0098	1.004	0.0800	0.9917	0.0592	0.0208	32	0.21	130.02	0
130	L-Lysine (%)	0226	1.004	0.0235	0.9917	0.0592	0.0208	32	0.21	130.00	0
130	L-Lysine (%)	0571	1.015	0.0530	0.9917	0.0592	0.0208	32	0.39	130.00	0
130	L-Lysine (%)	0644	1.020	0.0000	0.9917	0.0592	0.0208	32	0.48	130.00	0
130	L-Lysine (%)	2146	1.020	0.0200	0.9917	0.0592	0.0208	32	0.48	130.99	0
130	L-Lysine (%)	0859	1.023	0.0166	0.9917	0.0592	0.0208	32	0.52	130.00	0
130	L-Lysine (%)	2059	1.024	0.0070	0.9917	0.0592	0.0208	32	0.54	130.00	0
130	L-Lysine (%)	2196	1.026	0.0000	0.9917	0.0592	0.0208	32	0.58	130.05	0
130	L-Lysine (%)	0870	1.026	0.0655	0.9917	0.0592	0.0208	32	0.59	130.00	0
130	L-Lysine (%)	2193	1.050	0.0460	0.9917	0.0592	0.0208	32	0.98	130.05	0
130	L-Lysine (%)	0619	1.055	0.0500	0.9917	0.0592	0.0208	32	1.07	130.00	0
130	L-Lysine (%)	0227	1.055	0.0300	0.9917	0.0592	0.0208	32	1.07	130.99	0
130	L-Lysine (%)	0504	1.060	0.0000	0.9917	0.0592	0.0208	32	1.15	130.00	0
130	L-Lysine (%)	0407	1.093	0.0010	0.9917	0.0592	0.0208	32	1.70	130.05	0
130	L-Lysine (%)	0675	1.105	0.0300	0.9917	0.0592	0.0208	32	1.91	130.00	0
130	L-Lysine (%)	0910	0.8550	0.2300	0.9917	0.0592	0.0208	32	-2.31	130.00	1
131	Methionine (%)	0407	0.1218	0.0035	0.3059	0.0280	0.0138	32	-6.58	131.05	0
131	Methionine (%)	0918	0.2650	0.0100	0.3059	0.0280	0.0138	32	-1.46	131.00	0
131	Methionine (%)	0148	0.2755	0.0390	0.3059	0.0280	0.0138	32	-1.09	131.05	0
131	Methionine (%)	0512	0.2760	0.0453	0.3059	0.0280	0.0138	32	-1.07	131.00	0
131	Methionine (%)	0354	0.2795	0.0010	0.3059	0.0280	0.0138	32	-0.94	131.00	0
131	Methionine (%)	0910	0.2800	0.0600	0.3059	0.0280	0.0138	32	-0.93	131.00	0
131	Methionine (%)	0939	0.2800	0.0000	0.3059	0.0280	0.0138	32	-0.93	131.00	0
131	Methionine (%)	0848	0.2850	0.0100	0.3059	0.0280	0.0138	32	-0.75	131.00	0
131	Methionine (%)	2144	0.2850	0.0100	0.3059	0.0280	0.0138	32	-0.75	131.99	0
131	Methionine (%)	0941	0.2890	0.0000	0.3059	0.0280	0.0138	32	-0.61	131.00	0
131	Methionine (%)	2319	0.2950	0.0100	0.3059	0.0280	0.0138	32	-0.39	131.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
131	Methionine (%)	0968	0.2950	0.0300	0.3059	0.0280	0.0138	32	-0.39	131.00	0
131	Methionine (%)	2146	0.2950	0.0300	0.3059	0.0280	0.0138	32	-0.39	131.99	0
131	Methionine (%)	0619	0.2970	0.0280	0.3059	0.0280	0.0138	32	-0.32	131.00	0
131	Methionine (%)	0098	0.2980	0.0120	0.3059	0.0280	0.0138	32	-0.28	131.02	0
131	Methionine (%)	0652	0.3050	0.0300	0.3059	0.0280	0.0138	32	-0.03	131.00	0
131	Methionine (%)	0859	0.3078	0.0056	0.3059	0.0280	0.0138	32	0.07	131.00	0
131	Methionine (%)	2059	0.3115	0.0010	0.3059	0.0280	0.0138	32	0.20	131.00	0
131	Methionine (%)	2339	0.3115	0.0050	0.3059	0.0280	0.0138	32	0.20	131.00	0
131	Methionine (%)	0571	0.3140	0.0040	0.3059	0.0280	0.0138	32	0.29	131.00	0
131	Methionine (%)	2246	0.3171	0.0210	0.3059	0.0280	0.0138	32	0.40	131.05	0
131	Methionine (%)	0870	0.3178	0.0358	0.3059	0.0280	0.0138	32	0.42	131.00	0
131	Methionine (%)	0684	0.3200	0.0000	0.3059	0.0280	0.0138	32	0.50	131.00	0
131	Methionine (%)	0226	0.3225	0.0064	0.3059	0.0280	0.0138	32	0.59	131.00	0
131	Methionine (%)	0504	0.3250	0.0100	0.3059	0.0280	0.0138	32	0.68	131.00	0
131	Methionine (%)	2196	0.3250	0.0000	0.3059	0.0280	0.0138	32	0.68	131.05	0
131	Methionine (%)	0227	0.3250	0.0100	0.3059	0.0280	0.0138	32	0.68	131.99	0
131	Methionine (%)	0644	0.3430	0.0000	0.3059	0.0280	0.0138	32	1.32	131.00	0
131	Methionine (%)	0353	0.3450	0.0100	0.3059	0.0280	0.0138	32	1.40	131.00	0
131	Methionine (%)	0675	0.3450	0.0100	0.3059	0.0280	0.0138	32	1.40	131.00	0
131	Methionine (%)	2188	0.3560	0.0000	0.3059	0.0280	0.0138	32	1.79	131.05	0
131	Methionine (%)	2193	0.4270	0.0040	0.3059	0.0280	0.0138	32	4.33	131.05	0
132	Phenylalanine (%)	0407	0.7580	0.0000	0.8758	0.0335	0.0109	28	-3.52	132.05	0
132	Phenylalanine (%)	2319	0.8200	0.0000	0.8758	0.0335	0.0109	28	-1.67	132.05	0
132	Phenylalanine (%)	0148	0.8275	0.0250	0.8758	0.0335	0.0109	28	-1.44	132.05	0
132	Phenylalanine (%)	0619	0.8355	0.0030	0.8758	0.0335	0.0109	28	-1.20	132.00	0
132	Phenylalanine (%)	2193	0.8395	0.0150	0.8758	0.0335	0.0109	28	-1.08	132.05	0
132	Phenylalanine (%)	0968	0.8450	0.0100	0.8758	0.0335	0.0109	28	-0.92	132.00	0
132	Phenylalanine (%)	2196	0.8450	0.0000	0.8758	0.0335	0.0109	28	-0.92	132.05	0
132	Phenylalanine (%)	0939	0.8650	0.0100	0.8758	0.0335	0.0109	28	-0.32	132.00	0
132	Phenylalanine (%)	0098	0.8695	0.0070	0.8758	0.0335	0.0109	28	-0.19	132.02	0
132	Phenylalanine (%)	0652	0.8700	0.0400	0.8758	0.0335	0.0109	28	-0.17	132.00	0
132	Phenylalanine (%)	0684	0.8700	0.0000	0.8758	0.0335	0.0109	28	-0.17	132.00	0
132	Phenylalanine (%)	2146	0.8700	0.0200	0.8758	0.0335	0.0109	28	-0.17	132.99	0
132	Phenylalanine (%)	0870	0.8771	0.0339	0.8758	0.0335	0.0109	28	0.04	132.00	0
132	Phenylalanine (%)	0571	0.8775	0.0030	0.8758	0.0335	0.0109	28	0.05	132.00	0
132	Phenylalanine (%)	2339	0.8775	0.0190	0.8758	0.0335	0.0109	28	0.05	132.00	0
132	Phenylalanine (%)	0226	0.8839	0.0025	0.8758	0.0335	0.0109	28	0.24	132.00	0
132	Phenylalanine (%)	0353	0.8850	0.0100	0.8758	0.0335	0.0109	28	0.28	132.00	0
132	Phenylalanine (%)	0675	0.8850	0.0100	0.8758	0.0335	0.0109	28	0.28	132.00	0
132	Phenylalanine (%)	2246	0.8854	0.0201	0.8758	0.0335	0.0109	28	0.29	132.05	0
132	Phenylalanine (%)	2059	0.8900	0.0080	0.8758	0.0335	0.0109	28	0.42	132.00	0
132	Phenylalanine (%)	0227	0.8900	0.0200	0.8758	0.0335	0.0109	28	0.42	132.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
132	Phenylalanine (%)	0941	0.8940	0.0000	0.8758	0.0335	0.0109	28	0.54	132.00	0
132	Phenylalanine (%)	0354	0.8945	0.0270	0.8758	0.0335	0.0109	28	0.56	132.00	0
132	Phenylalanine (%)	0644	0.9060	0.0000	0.8758	0.0335	0.0109	28	0.90	132.00	0
132	Phenylalanine (%)	0504	0.9100	0.0000	0.8758	0.0335	0.0109	28	1.02	132.00	0
132	Phenylalanine (%)	2188	0.9440	0.0000	0.8758	0.0335	0.0109	28	2.04	132.05	0
132	Phenylalanine (%)	0859	0.9531	0.0124	0.8758	0.0335	0.0109	28	2.31	132.00	0
132	Phenylalanine (%)	0918	0.9650	0.0100	0.8758	0.0335	0.0109	28	2.66	132.00	0
132	Phenylalanine (%)	0910	0.7300	0.1200	0.8758	0.0335	0.0109	28	-4.35	132.00	1
133	Proline (%)	0859	1.075	0.0183	1.255	0.0573	0.0158	27	-3.14	133.00	0
133	Proline (%)	0939	1.075	0.0100	1.255	0.0573	0.0158	27	-3.14	133.00	0
133	Proline (%)	0148	1.175	0.0100	1.255	0.0573	0.0158	27	-1.39	133.05	0
133	Proline (%)	0407	1.182	0.0040	1.255	0.0573	0.0158	27	-1.27	133.05	0
133	Proline (%)	0968	1.220	0.0200	1.255	0.0573	0.0158	27	-0.61	133.00	0
133	Proline (%)	0870	1.224	0.0478	1.255	0.0573	0.0158	27	-0.55	133.00	0
133	Proline (%)	0684	1.230	0.0200	1.255	0.0573	0.0158	27	-0.43	133.00	0
133	Proline (%)	0354	1.232	0.0570	1.255	0.0573	0.0158	27	-0.41	133.00	0
133	Proline (%)	2319	1.235	0.0100	1.255	0.0573	0.0158	27	-0.35	133.05	0
133	Proline (%)	2339	1.236	0.0040	1.255	0.0573	0.0158	27	-0.33	133.00	0
133	Proline (%)	2196	1.238	0.0000	1.255	0.0573	0.0158	27	-0.29	133.05	0
133	Proline (%)	0941	1.243	0.0000	1.255	0.0573	0.0158	27	-0.21	133.00	0
133	Proline (%)	2246	1.244	0.0163	1.255	0.0573	0.0158	27	-0.19	133.05	0
133	Proline (%)	0504	1.250	0.0200	1.255	0.0573	0.0158	27	-0.09	133.00	0
133	Proline (%)	2193	1.257	0.0080	1.255	0.0573	0.0158	27	0.04	133.05	0
133	Proline (%)	0652	1.260	0.0600	1.255	0.0573	0.0158	27	0.09	133.00	0
133	Proline (%)	0226	1.261	0.0131	1.255	0.0573	0.0158	27	0.11	133.00	0
133	Proline (%)	2059	1.267	0.0070	1.255	0.0573	0.0158	27	0.20	133.00	0
133	Proline (%)	0353	1.270	0.0200	1.255	0.0573	0.0158	27	0.26	133.00	0
133	Proline (%)	0644	1.282	0.0000	1.255	0.0573	0.0158	27	0.47	133.00	0
133	Proline (%)	0227	1.285	0.0300	1.255	0.0573	0.0158	27	0.52	133.99	0
133	Proline (%)	2146	1.285	0.0100	1.255	0.0573	0.0158	27	0.52	133.99	0
133	Proline (%)	0571	1.307	0.0100	1.255	0.0573	0.0158	27	0.91	133.00	0
133	Proline (%)	0675	1.400	0.0200	1.255	0.0573	0.0158	27	2.53	133.00	0
133	Proline (%)	2188	1.405	0.0000	1.255	0.0573	0.0158	27	2.62	133.05	0
133	Proline (%)	0619	1.590	0.0000	1.255	0.0573	0.0158	27	5.84	133.00	0
133	Proline (%)	0918	1.675	0.0100	1.255	0.0573	0.0158	27	7.33	133.00	0
133	Proline (%)	0910	1.085	0.2700	1.255	0.0573	0.0158	27	-2.96	133.00	1
134	Serine (%)	0407	0.6910	0.0040	0.8942	0.0569	0.0111	28	-3.57	134.05	0
134	Serine (%)	0098	0.7880	0.0040	0.8942	0.0569	0.0111	28	-1.87	134.02	0
134	Serine (%)	0504	0.7900	0.0200	0.8942	0.0569	0.0111	28	-1.83	134.00	0
134	Serine (%)	2196	0.8250	0.0000	0.8942	0.0569	0.0111	28	-1.22	134.05	0
134	Serine (%)	0148	0.8450	0.0200	0.8942	0.0569	0.0111	28	-0.86	134.05	0
134	Serine (%)	0644	0.8590	0.0000	0.8942	0.0569	0.0111	28	-0.62	134.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
134	Serine (%)	0918	0.8600	0.0000	0.8942	0.0569	0.0111	28	-0.60	134.00	0
134	Serine (%)	0675	0.8650	0.0100	0.8942	0.0569	0.0111	28	-0.51	134.00	0
134	Serine (%)	0968	0.8650	0.0100	0.8942	0.0569	0.0111	28	-0.51	134.00	0
134	Serine (%)	2188	0.8690	0.0000	0.8942	0.0569	0.0111	28	-0.44	134.05	0
134	Serine (%)	2339	0.8765	0.0110	0.8942	0.0569	0.0111	28	-0.31	134.00	0
134	Serine (%)	0684	0.8800	0.0200	0.8942	0.0569	0.0111	28	-0.25	134.00	0
134	Serine (%)	2193	0.8800	0.0060	0.8942	0.0569	0.0111	28	-0.25	134.05	0
134	Serine (%)	2059	0.8840	0.0020	0.8942	0.0569	0.0111	28	-0.18	134.00	0
134	Serine (%)	0939	0.9000	0.0000	0.8942	0.0569	0.0111	28	0.10	134.00	0
134	Serine (%)	0941	0.9040	0.0000	0.8942	0.0569	0.0111	28	0.17	134.00	0
134	Serine (%)	0619	0.9060	0.0060	0.8942	0.0569	0.0111	28	0.21	134.00	0
134	Serine (%)	0227	0.9100	0.0200	0.8942	0.0569	0.0111	28	0.28	134.99	0
134	Serine (%)	0354	0.9255	0.0510	0.8942	0.0569	0.0111	28	0.55	134.00	0
134	Serine (%)	0571	0.9265	0.0070	0.8942	0.0569	0.0111	28	0.57	134.00	0
134	Serine (%)	0353	0.9350	0.0100	0.8942	0.0569	0.0111	28	0.72	134.00	0
134	Serine (%)	0652	0.9350	0.0100	0.8942	0.0569	0.0111	28	0.72	134.00	0
134	Serine (%)	0226	0.9371	0.0190	0.8942	0.0569	0.0111	28	0.75	134.00	0
134	Serine (%)	2319	0.9500	0.0000	0.8942	0.0569	0.0111	28	0.98	134.05	0
134	Serine (%)	2146	0.9500	0.0200	0.8942	0.0569	0.0111	28	0.98	134.99	0
134	Serine (%)	0859	0.9636	0.0249	0.8942	0.0569	0.0111	28	1.22	134.00	0
134	Serine (%)	2246	0.9994	0.0121	0.8942	0.0569	0.0111	28	1.85	134.05	0
134	Serine (%)	0870	1.001	0.0236	0.8942	0.0569	0.0111	28	1.88	134.00	0
134	Serine (%)	0910	0.7800	0.1400	0.8942	0.0569	0.0111	28	-2.01	134.00	1
135	Threonine (%)	0407	0.5770	0.0000	0.7165	0.0205	0.0118	28	-6.82	135.05	0
135	Threonine (%)	2196	0.6280	0.0000	0.7165	0.0205	0.0118	28	-4.33	135.05	0
135	Threonine (%)	0939	0.6850	0.0100	0.7165	0.0205	0.0118	28	-1.54	135.00	0
135	Threonine (%)	0226	0.6968	0.0101	0.7165	0.0205	0.0118	28	-0.97	135.00	0
135	Threonine (%)	2319	0.7000	0.0200	0.7165	0.0205	0.0118	28	-0.81	135.05	0
135	Threonine (%)	0148	0.7025	0.0750	0.7165	0.0205	0.0118	28	-0.69	135.05	0
135	Threonine (%)	2059	0.7040	0.0100	0.7165	0.0205	0.0118	28	-0.61	135.00	0
135	Threonine (%)	0098	0.7040	0.0080	0.7165	0.0205	0.0118	28	-0.61	135.02	0
135	Threonine (%)	0504	0.7050	0.0100	0.7165	0.0205	0.0118	28	-0.56	135.00	0
135	Threonine (%)	2339	0.7060	0.0040	0.7165	0.0205	0.0118	28	-0.52	135.00	0
135	Threonine (%)	2193	0.7100	0.0100	0.7165	0.0205	0.0118	28	-0.32	135.05	0
135	Threonine (%)	0644	0.7110	0.0000	0.7165	0.0205	0.0118	28	-0.27	135.00	0
135	Threonine (%)	0941	0.7140	0.0000	0.7165	0.0205	0.0118	28	-0.12	135.00	0
135	Threonine (%)	0918	0.7150	0.0100	0.7165	0.0205	0.0118	28	-0.08	135.00	0
135	Threonine (%)	0968	0.7150	0.0100	0.7165	0.0205	0.0118	28	-0.08	135.00	0
135	Threonine (%)	0684	0.7200	0.0200	0.7165	0.0205	0.0118	28	0.17	135.00	0
135	Threonine (%)	2188	0.7200	0.0000	0.7165	0.0205	0.0118	28	0.17	135.05	0
135	Threonine (%)	0619	0.7250	0.0100	0.7165	0.0205	0.0118	28	0.41	135.00	0
135	Threonine (%)	0675	0.7250	0.0100	0.7165	0.0205	0.0118	28	0.41	135.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
135	Threonine (%)	0227	0.7250	0.0100	0.7165	0.0205	0.0118	28	0.41	135.99	0
135	Threonine (%)	0652	0.7300	0.0000	0.7165	0.0205	0.0118	28	0.66	135.00	0
135	Threonine (%)	2146	0.7300	0.0200	0.7165	0.0205	0.0118	28	0.66	135.99	0
135	Threonine (%)	0571	0.7345	0.0050	0.7165	0.0205	0.0118	28	0.88	135.00	0
135	Threonine (%)	0353	0.7350	0.0100	0.7165	0.0205	0.0118	28	0.90	135.00	0
135	Threonine (%)	0354	0.7375	0.0350	0.7165	0.0205	0.0118	28	1.02	135.00	0
135	Threonine (%)	0859	0.7461	0.0111	0.7165	0.0205	0.0118	28	1.44	135.00	0
135	Threonine (%)	0870	0.7624	0.0211	0.7165	0.0205	0.0118	28	2.24	135.00	0
135	Threonine (%)	2246	0.7849	0.0004	0.7165	0.0205	0.0118	28	3.34	135.05	0
135	Threonine (%)	0910	0.6200	0.1200	0.7165	0.0205	0.0118	28	-4.72	135.00	1
136	Tryptophan (%)	2246	0.1311	0.0044	0.2268	0.0257	0.0100	21	-3.73	136.05	0
136	Tryptophan (%)	0910	0.1600	0.0400	0.2268	0.0257	0.0100	21	-2.60	136.00	0
136	Tryptophan (%)	2022	0.1950	0.0100	0.2268	0.0257	0.0100	21	-1.24	136.01	0
136	Tryptophan (%)	2196	0.1970	0.0000	0.2268	0.0257	0.0100	21	-1.16	136.03	0
136	Tryptophan (%)	2339	0.2160	0.0040	0.2268	0.0257	0.0100	21	-0.42	136.03	0
136	Tryptophan (%)	0098	0.2180	0.0000	0.2268	0.0257	0.0100	21	-0.34	136.02	0
136	Tryptophan (%)	0504	0.2200	0.0200	0.2268	0.0257	0.0100	21	-0.26	136.99	0
136	Tryptophan (%)	0644	0.2230	0.0000	0.2268	0.0257	0.0100	21	-0.15	136.01	0
136	Tryptophan (%)	2059	0.2245	0.0010	0.2268	0.0257	0.0100	21	-0.09	136.03	0
136	Tryptophan (%)	0571	0.2250	0.0020	0.2268	0.0257	0.0100	21	-0.07	136.03	0
136	Tryptophan (%)	0619	0.2255	0.0050	0.2268	0.0257	0.0100	21	-0.05	136.00	0
136	Tryptophan (%)	0941	0.2260	0.0000	0.2268	0.0257	0.0100	21	-0.03	136.00	0
136	Tryptophan (%)	0684	0.2350	0.0100	0.2268	0.0257	0.0100	21	0.32	136.00	0
136	Tryptophan (%)	0353	0.2350	0.0300	0.2268	0.0257	0.0100	21	0.32	136.03	0
136	Tryptophan (%)	0939	0.2350	0.0100	0.2268	0.0257	0.0100	21	0.32	136.03	0
136	Tryptophan (%)	2319	0.2350	0.0100	0.2268	0.0257	0.0100	21	0.32	136.05	0
136	Tryptophan (%)	0859	0.2354	0.0019	0.2268	0.0257	0.0100	21	0.33	136.03	0
136	Tryptophan (%)	0918	0.2450	0.0100	0.2268	0.0257	0.0100	21	0.71	136.00	0
136	Tryptophan (%)	0407	0.2685	0.0310	0.2268	0.0257	0.0100	21	1.62	136.05	0
136	Tryptophan (%)	0227	0.2700	0.0000	0.2268	0.0257	0.0100	21	1.68	136.00	0
136	Tryptophan (%)	0870	0.2929	0.0216	0.2268	0.0257	0.0100	21	2.57	136.00	0
137	Tyrosine (%)	0910	0.2650	0.0300	0.6003	0.0614	0.0171	25	-5.46	137.00	0
137	Tyrosine (%)	0407	0.4800	0.0020	0.6003	0.0614	0.0171	25	-1.96	137.05	0
137	Tyrosine (%)	0870	0.5195	0.0398	0.6003	0.0614	0.0171	25	-1.32	137.00	0
137	Tyrosine (%)	0675	0.5300	0.0200	0.6003	0.0614	0.0171	25	-1.15	137.00	0
137	Tyrosine (%)	0354	0.5385	0.0350	0.6003	0.0614	0.0171	25	-1.01	137.00	0
137	Tyrosine (%)	0227	0.5600	0.0400	0.6003	0.0614	0.0171	25	-0.66	137.99	0
137	Tyrosine (%)	0968	0.5650	0.0100	0.6003	0.0614	0.0171	25	-0.58	137.00	0
137	Tyrosine (%)	2319	0.5850	0.0300	0.6003	0.0614	0.0171	25	-0.25	137.05	0
137	Tyrosine (%)	2188	0.5890	0.0000	0.6003	0.0614	0.0171	25	-0.18	137.05	0
137	Tyrosine (%)	0684	0.5900	0.0000	0.6003	0.0614	0.0171	25	-0.17	137.00	0
137	Tyrosine (%)	0941	0.5930	0.0000	0.6003	0.0614	0.0171	25	-0.12	137.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
137	Tyrosine (%)	2246	0.5934	0.0231	0.6003	0.0614	0.0171	25	-0.11	137.05	0
137	Tyrosine (%)	0098	0.5980	0.0420	0.6003	0.0614	0.0171	25	-0.04	137.02	0
137	Tyrosine (%)	0504	0.6000	0.0400	0.6003	0.0614	0.0171	25	0.00	137.00	0
137	Tyrosine (%)	2146	0.6000	0.0200	0.6003	0.0614	0.0171	25	0.00	137.99	0
137	Tyrosine (%)	2196	0.6200	0.0000	0.6003	0.0614	0.0171	25	0.32	137.05	0
137	Tyrosine (%)	2193	0.6205	0.0010	0.6003	0.0614	0.0171	25	0.33	137.05	0
137	Tyrosine (%)	0226	0.6320	0.0108	0.6003	0.0614	0.0171	25	0.52	137.00	0
137	Tyrosine (%)	0939	0.6350	0.0100	0.6003	0.0614	0.0171	25	0.57	137.00	0
137	Tyrosine (%)	0644	0.6360	0.0000	0.6003	0.0614	0.0171	25	0.58	137.00	0
137	Tyrosine (%)	0353	0.6600	0.0000	0.6003	0.0614	0.0171	25	0.97	137.00	0
137	Tyrosine (%)	2339	0.6690	0.0040	0.6003	0.0614	0.0171	25	1.12	137.00	0
137	Tyrosine (%)	0148	0.6745	0.0270	0.6003	0.0614	0.0171	25	1.21	137.05	0
137	Tyrosine (%)	0918	0.6900	0.0400	0.6003	0.0614	0.0171	25	1.46	137.00	0
137	Tyrosine (%)	2059	0.6955	0.0030	0.6003	0.0614	0.0171	25	1.55	137.00	0
138	Valine (%)	2319	0.7150	0.0300	0.8983	0.0476	0.0152	28	-3.85	138.05	0
138	Valine (%)	2196	0.8180	0.0000	0.8983	0.0476	0.0152	28	-1.68	138.05	0
138	Valine (%)	0870	0.8400	0.0066	0.8983	0.0476	0.0152	28	-1.22	138.00	0
138	Valine (%)	2188	0.8410	0.0000	0.8983	0.0476	0.0152	28	-1.20	138.05	0
138	Valine (%)	0939	0.8450	0.0100	0.8983	0.0476	0.0152	28	-1.12	138.00	0
138	Valine (%)	0941	0.8550	0.0000	0.8983	0.0476	0.0152	28	-0.91	138.00	0
138	Valine (%)	0968	0.8750	0.0100	0.8983	0.0476	0.0152	28	-0.49	138.00	0
138	Valine (%)	0407	0.8750	0.0000	0.8983	0.0476	0.0152	28	-0.49	138.05	0
138	Valine (%)	2193	0.8805	0.0270	0.8983	0.0476	0.0152	28	-0.37	138.05	0
138	Valine (%)	0652	0.8850	0.0100	0.8983	0.0476	0.0152	28	-0.28	138.00	0
138	Valine (%)	0918	0.8850	0.0300	0.8983	0.0476	0.0152	28	-0.28	138.00	0
138	Valine (%)	0148	0.8935	0.0030	0.8983	0.0476	0.0152	28	-0.10	138.05	0
138	Valine (%)	0353	0.8950	0.0100	0.8983	0.0476	0.0152	28	-0.07	138.00	0
138	Valine (%)	2146	0.8950	0.0300	0.8983	0.0476	0.0152	28	-0.07	138.99	0
138	Valine (%)	2246	0.8966	0.0021	0.8983	0.0476	0.0152	28	-0.04	138.05	0
138	Valine (%)	0354	0.9050	0.0160	0.8983	0.0476	0.0152	28	0.14	138.00	0
138	Valine (%)	0859	0.9053	0.0153	0.8983	0.0476	0.0152	28	0.15	138.00	0
138	Valine (%)	0684	0.9100	0.0200	0.8983	0.0476	0.0152	28	0.25	138.00	0
138	Valine (%)	0619	0.9165	0.0590	0.8983	0.0476	0.0152	28	0.38	138.00	0
138	Valine (%)	0098	0.9200	0.0260	0.8983	0.0476	0.0152	28	0.46	138.02	0
138	Valine (%)	0226	0.9207	0.0162	0.8983	0.0476	0.0152	28	0.47	138.00	0
138	Valine (%)	0571	0.9210	0.0120	0.8983	0.0476	0.0152	28	0.48	138.00	0
138	Valine (%)	2059	0.9220	0.0140	0.8983	0.0476	0.0152	28	0.50	138.00	0
138	Valine (%)	2339	0.9415	0.0090	0.8983	0.0476	0.0152	28	0.91	138.00	0
138	Valine (%)	0227	0.9650	0.0100	0.8983	0.0476	0.0152	28	1.40	138.99	0
138	Valine (%)	0504	0.9700	0.0400	0.8983	0.0476	0.0152	28	1.51	138.00	0
138	Valine (%)	0644	0.9850	0.0000	0.8983	0.0476	0.0152	28	1.82	138.00	0
138	Valine (%)	0675	1.000	0.0200	0.8983	0.0476	0.0152	28	2.14	138.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
138	Valine (%)	0910	0.7850	0.2300	0.8983	0.0476	0.0152	28	-2.38	138.00	1
139	Taurine (%)	0407	0.0070	0.0000				2		139.05	0
139	Taurine (%)	0504	0.2400	0.0200				2		139.00	0
139	Taurine (%)	0098	< 0.01					2		139.02	5
139	Taurine (%)	0227	< 0.01					2		139.99	5
160	Fructose (%)	0619	0.1180	0.0040				3		160.99	0
160	Fructose (%)	0297	0.1450	0.0100				3		160.10	0
160	Fructose (%)	0227	0.3450	0.0100				3		160.99	0
161	Galactose (%)	0297	0.0000	0.0000				0		161.10	4
162	Glucose (%)	0619	0.1020	0.0040				3		162.99	0
162	Glucose (%)	0297	0.1600	0.0200				3		162.10	0
162	Glucose (%)	0227	0.2450	0.0100				3		162.99	0
162	Glucose (%)	0689	4.000	0.4000				3		162.99	2
163	Lactose (%)	0297	0.0000	0.0000				0		163.10	4
163	Lactose (%)	0619	0.0000	0.0000				0		163.99	4
163	Lactose (%)	0227	< 0.15					0		163.99	5
164	Maltose (%)	0297	0.2200	0.1600				1		164.10	0
164	Maltose (%)	0227	< 0.15					1		164.99	5
165	Sucrose (%)	0619	2.085	0.0500				3		165.99	0
165	Sucrose (%)	0297	2.250	0.2000				3		165.10	0
165	Sucrose (%)	0227	2.385	0.0300				3		165.99	0
166	Raffinose (%)	0227	0.3550	0.0100				3		166.99	0
166	Raffinose (%)	0619	0.3680	0.0100				3		166.99	0
166	Raffinose (%)	0297	0.4350	0.0100				3		166.10	0
167	Stachyose (%)	0619	0.8510	0.0300				3		167.99	0
167	Stachyose (%)	0227	1.010	0.0400				3		167.99	0
167	Stachyose (%)	0297	1.090	0.0200				3		167.10	0
345	Amprolium (ppm)	0148	40.06	3.287	97.51	5.868	1.501	10	-9.79	345.04	0
345	Amprolium (ppm)	0043	90.77	1.294	97.51	5.868	1.501	10	-1.15	345.00	0
345	Amprolium (ppm)	0218	94.60	0.2000	97.51	5.868	1.501	10	-0.50	345.02	0
345	Amprolium (ppm)	0014	96.95	2.100	97.51	5.868	1.501	10	-0.09	345.02	0
345	Amprolium (ppm)	0032	97.00	2.000	97.51	5.868	1.501	10	-0.09	345.02	0
345	Amprolium (ppm)	0047	99.26	1.290	97.51	5.868	1.501	10	0.30	345.99	0
345	Amprolium (ppm)	0033	99.30	3.600	97.51	5.868	1.501	10	0.31	345.02	0
345	Amprolium (ppm)	0964	99.57	0.0423	97.51	5.868	1.501	10	0.35	345.02	0
345	Amprolium (ppm)	0036	103.4	0.2000	97.51	5.868	1.501	10	1.00	345.00	0
345	Amprolium (ppm)	0001	105.5	1.000	97.51	5.868	1.501	10	1.36	345.02	0
345	Amprolium (ppm)	0013	105.1	17.80	97.51	5.868	1.501	10	1.29	345.00	1
348	Bacitracin (ppm)	0875	11.85	0.5000				3		348.06	0
348	Bacitracin (ppm)	0027	16.55	0.5500				3		348.00	0
348	Bacitracin (ppm)	0016	23.85	0.5000				3		348.01	0
351	Chlortetracycline (ppm)	0148	0.8320	0.0660				2		351.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
351	Chlortetracycline (ppm)	0218	1.441	0.1671				2		351.00	0
354	Decoquinatate (ppm)	0148	0.9925	0.0550				3		354.04	0
354	Decoquinatate (ppm)	2113	1.050	0.1000				3		354.04	0
354	Decoquinatate (ppm)	0218	1.400	0.0000				3		354.01	0
361	Lasalocid Sodium (ppm)	2113	1.250	0.1000				2		361.05	0
361	Lasalocid Sodium (ppm)	0148	2.825	0.5100				2		361.03	0
361	Lasalocid Sodium (ppm)	0014	< 1					2		361.03	5
365	Monensin (ppm)	2113	0.4300	0.0600				1		365.05	0
379	Salinomycin (ppm)	2113	< 0.25					0		379.05	5
382	Sulfamethazine (ppm)	0148	1.990	0.0200				2		382.04	0
382	Sulfamethazine (ppm)	0218	2.200	0.2000				2		382.02	0
391	Narasin (ppm)	2113	< 0.25					0		391.03	5
400	Water Activity (Units)	2303	0.4425	0.0110	0.4815	0.0295	0.0051	13	-1.32	400.01	0
400	Water Activity (Units)	0123	0.4550	0.0100	0.4815	0.0295	0.0051	13	-0.90	400.01	0
400	Water Activity (Units)	0942	0.4585	0.0170	0.4815	0.0295	0.0051	13	-0.78	400.01	0
400	Water Activity (Units)	2174	0.4650	0.0100	0.4815	0.0295	0.0051	13	-0.56	400.01	0
400	Water Activity (Units)	2268	0.4730	0.0040	0.4815	0.0295	0.0051	13	-0.29	400.01	0
400	Water Activity (Units)	0853	0.4744	0.0020	0.4815	0.0295	0.0051	13	-0.24	400.01	0
400	Water Activity (Units)	2168	0.4750	0.0040	0.4815	0.0295	0.0051	13	-0.22	400.01	0
400	Water Activity (Units)	2193	0.4810	0.0000	0.4815	0.0295	0.0051	13	-0.02	400.01	0
400	Water Activity (Units)	2073	0.4830	0.0000	0.4815	0.0295	0.0051	13	0.05	400.99	0
400	Water Activity (Units)	0407	0.4861	0.0026	0.4815	0.0295	0.0051	13	0.16	400.01	0
400	Water Activity (Units)	2181	0.5198	0.0038	0.4815	0.0295	0.0051	13	1.30	400.01	0
400	Water Activity (Units)	0885	0.5200	0.0000	0.4815	0.0295	0.0051	13	1.31	400.01	0
400	Water Activity (Units)	2109	0.5674	0.0018	0.4815	0.0295	0.0051	13	2.91	400.01	0
516	Arsenic, Total (As) (ppm)	0045	0.0610	0.0040	0.0973	0.0177	0.0079	8	-2.05	516.00	0
516	Arsenic, Total (As) (ppm)	0186	0.0840	0.0040	0.0973	0.0177	0.0079	8	-0.75	516.52	0
516	Arsenic, Total (As) (ppm)	0171	0.0925	0.0170	0.0973	0.0177	0.0079	8	-0.27	516.00	0
516	Arsenic, Total (As) (ppm)	0407	0.0933	0.0204	0.0973	0.0177	0.0079	8	-0.23	516.53	0
516	Arsenic, Total (As) (ppm)	0553	0.1020	0.0000	0.0973	0.0177	0.0079	8	0.26	516.53	0
516	Arsenic, Total (As) (ppm)	0910	0.1050	0.0100	0.0973	0.0177	0.0079	8	0.43	516.52	0
516	Arsenic, Total (As) (ppm)	0227	0.1110	0.0080	0.0973	0.0177	0.0079	8	0.77	516.53	0
516	Arsenic, Total (As) (ppm)	0939	0.1200	0.0000	0.0973	0.0177	0.0079	8	1.28	516.53	0
516	Arsenic, Total (As) (ppm)	0682	< 10		0.0973	0.0177	0.0079	8		516.43	5
518	Cadmium (ppm)	0171	0.0245	0.0070	0.0500	0.0057	0.0071	9	-4.50	518.41	0
518	Cadmium (ppm)	0939	0.0450	0.0100	0.0500	0.0057	0.0071	9	-0.89	518.53	0
518	Cadmium (ppm)	0407	0.0454	0.0113	0.0500	0.0057	0.0071	9	-0.83	518.43	0
518	Cadmium (ppm)	0910	0.0500	0.0000	0.0500	0.0057	0.0071	9	-0.01	518.52	0
518	Cadmium (ppm)	0407	0.0506	0.0047	0.0500	0.0057	0.0071	9	0.09	518.53	0
518	Cadmium (ppm)	0186	0.0535	0.0030	0.0500	0.0057	0.0071	9	0.61	518.52	0
518	Cadmium (ppm)	0227	0.0545	0.0130	0.0500	0.0057	0.0071	9	0.79	518.53	0
518	Cadmium (ppm)	0407	0.0546	0.0092	0.0500	0.0057	0.0071	9	0.81	518.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
518	Cadmium (ppm)	0553	0.0553	0.0053	0.0500	0.0057	0.0071	9	0.92	518.53	0
518	Cadmium (ppm)	0682	< 0.2		0.0500	0.0057	0.0071	9		518.43	5
520	Chromium, Total (Cr) (ppm)	0171	1.989	0.0200	3.446	1.308	0.1097	11	-1.11	520.41	0
520	Chromium, Total (Cr) (ppm)	0186	2.000	0.1200	3.446	1.308	0.1097	11	-1.11	520.52	0
520	Chromium, Total (Cr) (ppm)	0968	2.060	0.2240	3.446	1.308	0.1097	11	-1.06	520.43	0
520	Chromium, Total (Cr) (ppm)	0407	2.489	0.0130	3.446	1.308	0.1097	11	-0.73	520.41	0
520	Chromium, Total (Cr) (ppm)	0407	2.957	0.2819	3.446	1.308	0.1097	11	-0.37	520.43	0
520	Chromium, Total (Cr) (ppm)	0164	3.925	0.2700	3.446	1.308	0.1097	11	0.37	520.53	0
520	Chromium, Total (Cr) (ppm)	0553	4.230	0.0600	3.446	1.308	0.1097	11	0.60	520.53	0
520	Chromium, Total (Cr) (ppm)	0407	4.316	0.0176	3.446	1.308	0.1097	11	0.67	520.53	0
520	Chromium, Total (Cr) (ppm)	0510	4.410	0.1000	3.446	1.308	0.1097	11	0.74	520.43	0
520	Chromium, Total (Cr) (ppm)	0682	4.730	0.0000	3.446	1.308	0.1097	11	0.98	520.43	0
520	Chromium, Total (Cr) (ppm)	0045	4.800	0.1000	3.446	1.308	0.1097	11	1.04	520.42	0
520	Chromium, Total (Cr) (ppm)	0939	4.465	0.6900	3.446	1.308	0.1097	11	0.78	520.53	1
526	Lead (ppm)	0407	0.0517	0.1031	0.1093	0.0257	0.0375	8	-2.24	526.41	0
526	Lead (ppm)	0171	0.0955	0.0090	0.1093	0.0257	0.0375	8	-0.54	526.41	0
526	Lead (ppm)	0407	0.0991	0.0027	0.1093	0.0257	0.0375	8	-0.40	526.53	0
526	Lead (ppm)	0227	0.1025	0.0010	0.1093	0.0257	0.0375	8	-0.26	526.53	0
526	Lead (ppm)	0553	0.1140	0.0080	0.1093	0.0257	0.0375	8	0.18	526.53	0
526	Lead (ppm)	0186	0.1195	0.0050	0.1093	0.0257	0.0375	8	0.40	526.52	0
526	Lead (ppm)	0939	0.1250	0.0100	0.1093	0.0257	0.0375	8	0.61	526.53	0
526	Lead (ppm)	0407	0.1627	0.1613	0.1093	0.0257	0.0375	8	2.08	526.43	0
526	Lead (ppm)	0682	3.720	0.0000	0.1093	0.0257	0.0375	8	140.42	526.43	2
529	Mercury (ppb)	0407	0.5303	0.0353				1		529.99	0
529	Mercury (ppb)	0227	< 10					1		529.99	5
529	Mercury (ppb)	0682	< 10000					1		529.99	5
539	Nickel (ppm)	0171	1.935	0.0160	2.852	0.6658	0.0941	8	-1.38	539.41	0
539	Nickel (ppm)	0186	2.225	0.0900	2.852	0.6658	0.0941	8	-0.94	539.52	0
539	Nickel (ppm)	0407	2.650	0.0512	2.852	0.6658	0.0941	8	-0.30	539.41	0
539	Nickel (ppm)	0407	2.727	0.3115	2.852	0.6658	0.0941	8	-0.19	539.43	0
539	Nickel (ppm)	0407	2.945	0.0244	2.852	0.6658	0.0941	8	0.14	539.53	0
539	Nickel (ppm)	0939	3.320	0.2400	2.852	0.6658	0.0941	8	0.70	539.53	0
539	Nickel (ppm)	0553	3.360	0.0200	2.852	0.6658	0.0941	8	0.76	539.53	0
539	Nickel (ppm)	0682	3.650	0.0000	2.852	0.6658	0.0941	8	1.20	539.43	0
702	Butyric Acid (4:0) (%)	0619	0.0085	0.0010				1		702.00	0
704	Caproic Acid (6:0) (%)	0619	0.0005	0.0010				0		704.00	4
706	Caprylic acid (8:0) (%) (w/w))	0619	0.0000	0.0000				0		706.01	4
708	Capric acid (10:0) (%) (w/w))	0619	0.0010	0.0000				1		708.01	0
710	Lauric Acid (12:0) (%) (w/w))	0619	0.0020	0.0000				1		710.01	0
710	Lauric Acid (12:0) (%) (w/w))	0226	< 0.005					1		710.99	5
714	Myristic Acid (14:0) (%) (w/w))	0226	0.0169	0.0005				2		714.99	0
714	Myristic Acid (14:0) (%) (w/w))	0619	0.0200	0.0000				2		714.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
716	Palmitic Acid (16:0) (% (w/w))	0226	0.6446	0.0058				2	716.99	0	
716	Palmitic Acid (16:0) (% (w/w))	0619	0.7905	0.0070				2	716.01	0	
718	Palmitoleic Acid (9c-16:1) (% (w/w))	0226	0.0253	0.0005				1	718.99	0	
718	Palmitoleic Acid (9c-16:1) (% (w/w))	0619	0.0135	0.0270				1	718.01	4	
722	Stearic Acid (18:0) (% (w/w))	0226	0.2076	0.0064				2	722.99	0	
722	Stearic Acid (18:0) (% (w/w))	0619	0.2345	0.0050				2	722.01	0	
724	Oleic Acid (9c-18:1) (% (w/w))	0226	0.8742	0.0230				2	724.99	0	
724	Oleic Acid (9c-18:1) (% (w/w))	0619	1.030	0.0200				2	724.01	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0226	1.358	0.0568				2	726.99	0	
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0619	1.580	0.0200				2	726.01	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/w))	0226	0.0796	0.0043				2	728.99	0	
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/w))	0619	0.0970	0.0040				2	728.01	0	
730	Arachidic Acid (20:0) (% (w/w))	0226	0.0126	0.0000				2	730.99	0	
730	Arachidic Acid (20:0) (% (w/w))	0619	0.0130	0.0000				2	730.01	0	
732	Gondoic Acid (11c-20:1) (% (w/w))	0619	0.0160	0.0020				2	732.01	0	
732	Gondoic Acid (11c-20:1) (% (w/w))	0226	0.0172	0.0007				2	732.99	0	
736	Arachidonic Acid (5c,8c,11c,14c-20:4) (% (w/w))	0619	0.0060	0.0020				1	736.01	0	
738	Mead Acid (11c,14c,17c-20:3) (% (w/w))	0619	0.0000	0.0000				0	738.01	4	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c) (% (w/w))	0619	0.0000	0.0000				0	740.01	4	
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c) (% (w/w))	0226	< 0.005					0	740.99	5	
742	Behenic Acid (22:0) (% (w/w))	0226	0.0088	0.0003				1	742.99	0	
744	Erucic Acid (13c-22:1) (% (w/w))	0619	0.0000	0.0000				0	744.01	4	
744	Erucic Acid (13c-22:1) (% (w/w))	0226	< 0.005					0	744.99	5	
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c) (% (w/w))	0619	0.0000	0.0000				0	746.01	4	
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c) (% (w/w))	0226	< 0.005					0	746.99	5	
748	Lignoceric Acid (24:0) (% (w/w))	0226	0.0103	0.0004				1	748.99	0	
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22c) (% (w/w))	0619	0.0000	0.0000				0	750.01	4	
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22c) (% (w/w))	0226	< 0.005					0	750.99	5	
752	Nervonic Acid (24:1) isomers (% (w/w))	0619	0.0000	0.0000				0	752.01	4	
752	Nervonic Acid (24:1) isomers (% (w/w))	0226	< 0.005					0	752.99	5	
772	Total Fatty Acids (% (w/w))	0226	3.255	0.0839				1	772.99	0	

Notes: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2 and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = Rejected for duplicates too far apart, 2 = Rejected as outlier, 5 = A reporting limit and 4 = zeros submitted as values. Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). Flag 3 indicates not used in statistics.