



AAFCO
Proficiency Testing Program



Mycotoxin Contaminants Scheme

Swine Feed

Test Material Code # 202162

Analyte Proficiency Testing Report

Labs Reporting: 58

Analytes Reported 15

Issue Date : 05/31/2021

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
600	Total Aflatoxin (ppb)	2226	8.362	0.3260	23.23	7.856	4.236	46	-1.89	600.12	0
600	Total Aflatoxin (ppb)	2196	12.16	0.0000	23.23	7.856	4.236	46	-1.41	600.20	0
600	Total Aflatoxin (ppb)	2345	13.50	1.000	23.23	7.856	4.236	46	-1.24	600.14	0
600	Total Aflatoxin (ppb)	0619	15.65	3.100	23.23	7.856	4.236	46	-0.97	600.24	0
600	Total Aflatoxin (ppb)	0190	16.00	4.000	23.23	7.856	4.236	46	-0.92	600.14	0
600	Total Aflatoxin (ppb)	0682	16.80	0.0000	23.23	7.856	4.236	46	-0.82	600.24	0
600	Total Aflatoxin (ppb)	0010	17.50	3.000	23.23	7.856	4.236	46	-0.73	600.21	0
600	Total Aflatoxin (ppb)	2147	19.50	7.000	23.23	7.856	4.236	46	-0.48	600.04	0
600	Total Aflatoxin (ppb)	2033	19.55	4.900	23.23	7.856	4.236	46	-0.47	600.24	0
600	Total Aflatoxin (ppb)	2269	19.60	1.200	23.23	7.856	4.236	46	-0.46	600.98	0
600	Total Aflatoxin (ppb)	2192	19.78	0.1100	23.23	7.856	4.236	46	-0.44	600.20	0
600	Total Aflatoxin (ppb)	0208	19.80	1.200	23.23	7.856	4.236	46	-0.44	600.01	0
600	Total Aflatoxin (ppb)	2113	20.10	2.000	23.23	7.856	4.236	46	-0.40	600.01	0
600	Total Aflatoxin (ppb)	0353	20.29	0.4700	23.23	7.856	4.236	46	-0.38	600.24	0
600	Total Aflatoxin (ppb)	0027	20.41	0.3200	23.23	7.856	4.236	46	-0.36	600.21	0
600	Total Aflatoxin (ppb)	0051	21.40	3.600	23.23	7.856	4.236	46	-0.23	600.20	0
600	Total Aflatoxin (ppb)	2096	21.50	4.800	23.23	7.856	4.236	46	-0.22	600.01	0
600	Total Aflatoxin (ppb)	0004	21.50	1.000	23.23	7.856	4.236	46	-0.22	600.14	0
600	Total Aflatoxin (ppb)	0520	21.50	1.000	23.23	7.856	4.236	46	-0.22	600.14	0
600	Total Aflatoxin (ppb)	2169	21.50	0.4000	23.23	7.856	4.236	46	-0.22	600.98	0
600	Total Aflatoxin (ppb)	0511	21.55	0.5000	23.23	7.856	4.236	46	-0.21	600.01	0
600	Total Aflatoxin (ppb)	0202	21.80	2.800	23.23	7.856	4.236	46	-0.18	600.12	0
600	Total Aflatoxin (ppb)	0876	22.20	2.200	23.23	7.856	4.236	46	-0.13	600.01	0
600	Total Aflatoxin (ppb)	0160	22.45	1.700	23.23	7.856	4.236	46	-0.10	600.24	0
600	Total Aflatoxin (ppb)	0098	22.85	3.000	23.23	7.856	4.236	46	-0.05	600.22	0
600	Total Aflatoxin (ppb)	0529	23.55	1.100	23.23	7.856	4.236	46	0.04	600.01	0
600	Total Aflatoxin (ppb)	0034	23.68	0.6500	23.23	7.856	4.236	46	0.06	600.01	0
600	Total Aflatoxin (ppb)	0035	23.99	0.7200	23.23	7.856	4.236	46	0.10	600.21	0
600	Total Aflatoxin (ppb)	0563	24.00	2.000	23.23	7.856	4.236	46	0.10	600.20	0
600	Total Aflatoxin (ppb)	2302	24.14	0.3950	23.23	7.856	4.236	46	0.12	600.99	0
600	Total Aflatoxin (ppb)	0142	25.40	3.000	23.23	7.856	4.236	46	0.28	600.98	0
600	Total Aflatoxin (ppb)	2022	25.94	0.5100	23.23	7.856	4.236	46	0.34	600.08	0
600	Total Aflatoxin (ppb)	0015	27.21	6.830	23.23	7.856	4.236	46	0.51	600.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
600	Total Aflatoxin (ppb)	0218	27.55	0.1000	23.23	7.856	4.236	46	0.55	600.24	0
600	Total Aflatoxin (ppb)	0066	27.60	1.200	23.23	7.856	4.236	46	0.56	600.13	0
600	Total Aflatoxin (ppb)	0554	27.90	5.000	23.23	7.856	4.236	46	0.59	600.01	0
600	Total Aflatoxin (ppb)	0590	28.25	3.100	23.23	7.856	4.236	46	0.64	600.12	0
600	Total Aflatoxin (ppb)	0032	28.35	4.700	23.23	7.856	4.236	46	0.65	600.24	0
600	Total Aflatoxin (ppb)	0870	28.85	4.500	23.23	7.856	4.236	46	0.72	600.98	0
600	Total Aflatoxin (ppb)	0407	29.41	6.357	23.23	7.856	4.236	46	0.79	600.13	0
600	Total Aflatoxin (ppb)	2098	29.70	1.000	23.23	7.856	4.236	46	0.82	600.01	0
600	Total Aflatoxin (ppb)	0042	29.95	6.300	23.23	7.856	4.236	46	0.86	600.20	0
600	Total Aflatoxin (ppb)	0035	30.63	0.8900	23.23	7.856	4.236	46	0.94	600.13	0
600	Total Aflatoxin (ppb)	2129	33.85	0.9000	23.23	7.856	4.236	46	1.35	600.24	0
600	Total Aflatoxin (ppb)	0227	43.00	6.000	23.23	7.856	4.236	46	2.52	600.01	0
600	Total Aflatoxin (ppb)	0297	2,585	90.00	23.23	7.856	4.236	46	326.10	600.01	0
600	Total Aflatoxin (ppb)	0047	18,900	200.0	23.23	7.856	4.236	46	2402.89	600.02	2
601	Aflatoxin B1 (ppb)	2196	11.02	0.0000	20.37	6.913	1.727	28	-1.35	601.20	0
601	Aflatoxin B1 (ppb)	2207	11.50	3.000	20.37	6.913	1.727	28	-1.28	601.23	0
601	Aflatoxin B1 (ppb)	0682	15.10	0.0000	20.37	6.913	1.727	28	-0.76	601.23	0
601	Aflatoxin B1 (ppb)	0619	15.65	3.100	20.37	6.913	1.727	28	-0.68	601.23	0
601	Aflatoxin B1 (ppb)	0870	15.92	1.440	20.37	6.913	1.727	28	-0.64	601.23	0
601	Aflatoxin B1 (ppb)	2136	16.05	4.900	20.37	6.913	1.727	28	-0.63	601.23	0
601	Aflatoxin B1 (ppb)	2192	16.28	0.1700	20.37	6.913	1.727	28	-0.59	601.20	0
601	Aflatoxin B1 (ppb)	0010	16.50	2.800	20.37	6.913	1.727	28	-0.56	601.21	0
601	Aflatoxin B1 (ppb)	0028	16.66	0.0120	20.37	6.913	1.727	28	-0.54	601.23	0
601	Aflatoxin B1 (ppb)	2033	17.70	4.400	20.37	6.913	1.727	28	-0.39	601.23	0
601	Aflatoxin B1 (ppb)	0353	18.03	0.0500	20.37	6.913	1.727	28	-0.34	601.23	0
601	Aflatoxin B1 (ppb)	0027	18.32	0.0910	20.37	6.913	1.727	28	-0.30	601.21	0
601	Aflatoxin B1 (ppb)	0913	19.99	4.220	20.37	6.913	1.727	28	-0.06	601.23	0
601	Aflatoxin B1 (ppb)	0098	20.47	2.520	20.37	6.913	1.727	28	0.01	601.24	0
601	Aflatoxin B1 (ppb)	0160	20.60	1.600	20.37	6.913	1.727	28	0.03	601.23	0
601	Aflatoxin B1 (ppb)	2269	21.85	0.7000	20.37	6.913	1.727	28	0.21	601.98	0
601	Aflatoxin B1 (ppb)	0035	21.88	0.6300	20.37	6.913	1.727	28	0.22	601.21	0
601	Aflatoxin B1 (ppb)	0918	22.40	3.200	20.37	6.913	1.727	28	0.29	601.23	0
601	Aflatoxin B1 (ppb)	0563	22.50	1.000	20.37	6.913	1.727	28	0.31	601.20	0
601	Aflatoxin B1 (ppb)	2053	23.01	0.4000	20.37	6.913	1.727	28	0.38	601.23	0
601	Aflatoxin B1 (ppb)	0964	23.18	1.600	20.37	6.913	1.727	28	0.41	601.23	0
601	Aflatoxin B1 (ppb)	0910	23.35	0.2400	20.37	6.913	1.727	28	0.43	601.21	0
601	Aflatoxin B1 (ppb)	0218	24.85	0.1000	20.37	6.913	1.727	28	0.65	601.23	0
601	Aflatoxin B1 (ppb)	2022	25.08	0.2200	20.37	6.913	1.727	28	0.68	601.91	0
601	Aflatoxin B1 (ppb)	2140	25.50	1.000	20.37	6.913	1.727	28	0.74	601.23	0
601	Aflatoxin B1 (ppb)	0032	28.35	4.700	20.37	6.913	1.727	28	1.15	601.23	0
601	Aflatoxin B1 (ppb)	2302	28.50	3.000	20.37	6.913	1.727	28	1.18	601.99	0
601	Aflatoxin B1 (ppb)	2327	30.01	3.249	20.37	6.913	1.727	28	1.39	601.23	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
602	Aflatoxin B2 (ppb)	0913	0.8250	0.4300	1.904	0.6891	0.2649	21	-1.57	602.23	0
602	Aflatoxin B2 (ppb)	0010	0.8500	0.1000	1.904	0.6891	0.2649	21	-1.53	602.21	0
602	Aflatoxin B2 (ppb)	2196	1.143	0.0000	1.904	0.6891	0.2649	21	-1.10	602.20	0
602	Aflatoxin B2 (ppb)	0563	1.500	1.000	1.904	0.6891	0.2649	21	-0.59	602.20	0
602	Aflatoxin B2 (ppb)	0353	1.500	0.0600	1.904	0.6891	0.2649	21	-0.59	602.23	0
602	Aflatoxin B2 (ppb)	2192	1.660	0.0400	1.904	0.6891	0.2649	21	-0.35	602.20	0
602	Aflatoxin B2 (ppb)	0682	1.700	0.0000	1.904	0.6891	0.2649	21	-0.30	602.23	0
602	Aflatoxin B2 (ppb)	0870	1.770	0.1000	1.904	0.6891	0.2649	21	-0.19	602.23	0
602	Aflatoxin B2 (ppb)	2053	1.825	0.0700	1.904	0.6891	0.2649	21	-0.12	602.23	0
602	Aflatoxin B2 (ppb)	0160	1.835	0.1300	1.904	0.6891	0.2649	21	-0.10	602.23	0
602	Aflatoxin B2 (ppb)	2033	1.850	0.5000	1.904	0.6891	0.2649	21	-0.08	602.23	0
602	Aflatoxin B2 (ppb)	2136	1.850	0.5000	1.904	0.6891	0.2649	21	-0.08	602.23	0
602	Aflatoxin B2 (ppb)	0910	1.880	0.3000	1.904	0.6891	0.2649	21	-0.04	602.21	0
602	Aflatoxin B2 (ppb)	0918	2.000	0.6000	1.904	0.6891	0.2649	21	0.14	602.23	0
602	Aflatoxin B2 (ppb)	0027	2.093	0.4110	1.904	0.6891	0.2649	21	0.27	602.21	0
602	Aflatoxin B2 (ppb)	0035	2.115	0.0900	1.904	0.6891	0.2649	21	0.31	602.21	0
602	Aflatoxin B2 (ppb)	0098	2.380	0.4800	1.904	0.6891	0.2649	21	0.69	602.24	0
602	Aflatoxin B2 (ppb)	2327	2.573	0.1050	1.904	0.6891	0.2649	21	0.97	602.23	0
602	Aflatoxin B2 (ppb)	0218	2.700	0.0000	1.904	0.6891	0.2649	21	1.15	602.23	0
602	Aflatoxin B2 (ppb)	0964	3.274	0.0330	1.904	0.6891	0.2649	21	1.99	602.23	0
602	Aflatoxin B2 (ppb)	0028	4.132	0.6130	1.904	0.6891	0.2649	21	3.23	602.23	0
602	Aflatoxin B2 (ppb)	0032	0.0000	0.0000	1.904	0.6891	0.2649	21		602.23	4
602	Aflatoxin B2 (ppb)	0619	0.0000	0.0000	1.904	0.6891	0.2649	21		602.23	4
602	Aflatoxin B2 (ppb)	2207	< 1		1.904	0.6891	0.2649	21		602.23	5
602	Aflatoxin B2 (ppb)	2140	< 20		1.904	0.6891	0.2649	21		602.23	5
603	Aflatoxin G1 (ppb)	0913	0.4000	0.1000				3		603.23	0
603	Aflatoxin G1 (ppb)	0910	0.5100	0.1600				3		603.21	0
603	Aflatoxin G1 (ppb)	2302	4.000	0.0000				3		603.99	0
603	Aflatoxin G1 (ppb)	0032	0.0000	0.0000				3		603.23	4
603	Aflatoxin G1 (ppb)	0160	0.0000	0.0000				3		603.23	4
603	Aflatoxin G1 (ppb)	0619	0.0000	0.0000				3		603.23	4
603	Aflatoxin G1 (ppb)	2053	< 0.32					3		603.23	5
603	Aflatoxin G1 (ppb)	0682	< 0.5					3		603.23	5
603	Aflatoxin G1 (ppb)	0870	< 0.6					3		603.23	5
603	Aflatoxin G1 (ppb)	2136	< 0.83					3		603.23	5
603	Aflatoxin G1 (ppb)	0218	< 1					3		603.23	5
603	Aflatoxin G1 (ppb)	0918	< 1					3		603.23	5
603	Aflatoxin G1 (ppb)	0964	< 1					3		603.23	5
603	Aflatoxin G1 (ppb)	2207	< 1					3		603.23	5
603	Aflatoxin G1 (ppb)	2327	< 1.1					3		603.23	5
603	Aflatoxin G1 (ppb)	2033	< 5					3		603.23	5
603	Aflatoxin G1 (ppb)	0098	< 15					3		603.24	5

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
603	Aflatoxin G1 (ppb)	2140	< 20					3	603.23	5	
604	Aflatoxin G2 (ppb)	0032	0.0000	0.0000				0	604.23	4	
604	Aflatoxin G2 (ppb)	0160	0.0000	0.0000				0	604.23	4	
604	Aflatoxin G2 (ppb)	0619	0.0000	0.0000				0	604.23	4	
604	Aflatoxin G2 (ppb)	0910	0.0200	0.0400				0	604.21	4	
604	Aflatoxin G2 (ppb)	2053	< 0.49					0	604.23	5	
604	Aflatoxin G2 (ppb)	0682	< 0.5					0	604.23	5	
604	Aflatoxin G2 (ppb)	2136	< 0.83					0	604.23	5	
604	Aflatoxin G2 (ppb)	0218	< 1					0	604.23	5	
604	Aflatoxin G2 (ppb)	0870	< 1					0	604.23	5	
604	Aflatoxin G2 (ppb)	0913	< 1					0	604.23	5	
604	Aflatoxin G2 (ppb)	0918	< 1					0	604.23	5	
604	Aflatoxin G2 (ppb)	0964	< 1					0	604.23	5	
604	Aflatoxin G2 (ppb)	2207	< 1					0	604.23	5	
604	Aflatoxin G2 (ppb)	2327	< 1.6					0	604.23	5	
604	Aflatoxin G2 (ppb)	2033	< 5					0	604.23	5	
604	Aflatoxin G2 (ppb)	0098	< 5					0	604.24	5	
604	Aflatoxin G2 (ppb)	2140	< 20					0	604.23	5	
610	Deoxynivalenol (ppb)	0529	1.250	0.1000	1,475	445.7	120.2	53	-3.31	610.01	0
610	Deoxynivalenol (ppb)	0910	1.375	0.0100	1,475	445.7	120.2	53	-3.31	610.20	0
610	Deoxynivalenol (ppb)	2066	558.5	39.00	1,475	445.7	120.2	53	-2.06	610.07	0
610	Deoxynivalenol (ppb)	0066	915.0	30.00	1,475	445.7	120.2	53	-1.26	610.08	0
610	Deoxynivalenol (ppb)	2207	940.0	80.00	1,475	445.7	120.2	53	-1.20	610.23	0
610	Deoxynivalenol (ppb)	0563	970.5	183.0	1,475	445.7	120.2	53	-1.13	610.25	0
610	Deoxynivalenol (ppb)	2136	975.5	69.00	1,475	445.7	120.2	53	-1.12	610.23	0
610	Deoxynivalenol (ppb)	0032	1,125	10.00	1,475	445.7	120.2	53	-0.79	610.23	0
610	Deoxynivalenol (ppb)	0913	1,128	55.00	1,475	445.7	120.2	53	-0.78	610.23	0
610	Deoxynivalenol (ppb)	0028	1,143	121.1	1,475	445.7	120.2	53	-0.75	610.23	0
610	Deoxynivalenol (ppb)	0098	1,150	100.0	1,475	445.7	120.2	53	-0.73	610.20	0
610	Deoxynivalenol (ppb)	2140	1,195	80.00	1,475	445.7	120.2	53	-0.63	610.23	0
610	Deoxynivalenol (ppb)	0051	1,252	22.00	1,475	445.7	120.2	53	-0.50	610.20	0
610	Deoxynivalenol (ppb)	0353	1,266	14.00	1,475	445.7	120.2	53	-0.47	610.23	0
610	Deoxynivalenol (ppb)	2196	1,307	0.0000	1,475	445.7	120.2	53	-0.38	610.20	0
610	Deoxynivalenol (ppb)	0619	1,320	140.0	1,475	445.7	120.2	53	-0.35	610.23	0
610	Deoxynivalenol (ppb)	2129	1,328	123.0	1,475	445.7	120.2	53	-0.33	610.23	0
610	Deoxynivalenol (ppb)	2033	1,329	12.00	1,475	445.7	120.2	53	-0.33	610.23	0
610	Deoxynivalenol (ppb)	0202	1,349	101.0	1,475	445.7	120.2	53	-0.28	610.06	0
610	Deoxynivalenol (ppb)	0870	1,398	49.90	1,475	445.7	120.2	53	-0.17	610.99	0
610	Deoxynivalenol (ppb)	0511	1,400	0.0000	1,475	445.7	120.2	53	-0.17	610.01	0
610	Deoxynivalenol (ppb)	0218	1,420	80.00	1,475	445.7	120.2	53	-0.12	610.23	0
610	Deoxynivalenol (ppb)	0682	1,420	0.0000	1,475	445.7	120.2	53	-0.12	610.23	0
610	Deoxynivalenol (ppb)	0010	1,435	110.0	1,475	445.7	120.2	53	-0.09	610.23	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
610	Deoxynivalenol (ppb)	2053	1,441	28.00	1,475	445.7	120.2	53	-0.08	610.23	0
610	Deoxynivalenol (ppb)	0015	1,469	351.0	1,475	445.7	120.2	53	-0.01	610.99	0
610	Deoxynivalenol (ppb)	0870	1,470	20.00	1,475	445.7	120.2	53	-0.01	610.23	0
610	Deoxynivalenol (ppb)	0554	1,500	800.0	1,475	445.7	120.2	53	0.06	610.01	0
610	Deoxynivalenol (ppb)	0918	1,505	34.43	1,475	445.7	120.2	53	0.07	610.23	0
610	Deoxynivalenol (ppb)	0413	1,550	100.0	1,475	445.7	120.2	53	0.17	610.01	0
610	Deoxynivalenol (ppb)	2022	1,560	120.0	1,475	445.7	120.2	53	0.19	610.05	0
610	Deoxynivalenol (ppb)	0590	1,560	20.00	1,475	445.7	120.2	53	0.19	610.07	0
610	Deoxynivalenol (ppb)	2098	1,570	40.00	1,475	445.7	120.2	53	0.21	610.01	0
610	Deoxynivalenol (ppb)	0047	1,580	220.0	1,475	445.7	120.2	53	0.24	610.02	0
610	Deoxynivalenol (ppb)	0407	1,582	5.000	1,475	445.7	120.2	53	0.24	610.07	0
610	Deoxynivalenol (ppb)	2113	1,600	200.0	1,475	445.7	120.2	53	0.28	610.01	0
610	Deoxynivalenol (ppb)	0413	1,600	0.0000	1,475	445.7	120.2	53	0.28	610.20	0
610	Deoxynivalenol (ppb)	0035	1,616	189.0	1,475	445.7	120.2	53	0.32	610.08	0
610	Deoxynivalenol (ppb)	0027	1,700	400.0	1,475	445.7	120.2	53	0.50	610.01	0
610	Deoxynivalenol (ppb)	0142	1,700	200.0	1,475	445.7	120.2	53	0.50	610.98	0
610	Deoxynivalenol (ppb)	0160	1,707	116.0	1,475	445.7	120.2	53	0.52	610.23	0
610	Deoxynivalenol (ppb)	2192	1,720	40.00	1,475	445.7	120.2	53	0.55	610.01	0
610	Deoxynivalenol (ppb)	0208	1,755	330.0	1,475	445.7	120.2	53	0.63	610.01	0
610	Deoxynivalenol (ppb)	0297	1,800	200.0	1,475	445.7	120.2	53	0.73	610.01	0
610	Deoxynivalenol (ppb)	0042	1,850	100.0	1,475	445.7	120.2	53	0.84	610.01	0
610	Deoxynivalenol (ppb)	0964	1,887	134.1	1,475	445.7	120.2	53	0.93	610.23	0
610	Deoxynivalenol (ppb)	0227	1,900	0.0000	1,475	445.7	120.2	53	0.95	610.01	0
610	Deoxynivalenol (ppb)	2302	1,926	40.75	1,475	445.7	120.2	53	1.01	610.99	0
610	Deoxynivalenol (ppb)	0870	2,000	0.0000	1,475	445.7	120.2	53	1.18	610.01	0
610	Deoxynivalenol (ppb)	2169	2,000	400.0	1,475	445.7	120.2	53	1.18	610.98	0
610	Deoxynivalenol (ppb)	2327	2,084	413.3	1,475	445.7	120.2	53	1.37	610.23	0
610	Deoxynivalenol (ppb)	2147	2,100	400.0	1,475	445.7	120.2	53	1.40	610.04	0
610	Deoxynivalenol (ppb)	2269	2,975	50.00	1,475	445.7	120.2	53	3.37	610.98	0
620	Total Fumonisin (ppb)	0529	4,950	0.1000	5,823	1,695	512.7	33	-3.43	620.02	0
620	Total Fumonisin (ppb)	0098	3,250	300.0	5,823	1,695	512.7	33	-1.52	620.11	0
620	Total Fumonisin (ppb)	0190	3,500	400.0	5,823	1,695	512.7	33	-1.37	620.11	0
620	Total Fumonisin (ppb)	2098	4,615	650.0	5,823	1,695	512.7	33	-0.71	620.02	0
620	Total Fumonisin (ppb)	0554	4,700	200.0	5,823	1,695	512.7	33	-0.66	620.01	0
620	Total Fumonisin (ppb)	0619	4,882	286.0	5,823	1,695	512.7	33	-0.56	620.23	0
620	Total Fumonisin (ppb)	2066	5,065	410.0	5,823	1,695	512.7	33	-0.45	620.09	0
620	Total Fumonisin (ppb)	2147	5,100	1,000	5,823	1,695	512.7	33	-0.43	620.04	0
620	Total Fumonisin (ppb)	0047	5,131	302.0	5,823	1,695	512.7	33	-0.41	620.02	0
620	Total Fumonisin (ppb)	0010	5,200	0.0000	5,823	1,695	512.7	33	-0.37	620.23	0
620	Total Fumonisin (ppb)	0042	5,350	100.0	5,823	1,695	512.7	33	-0.28	620.01	0
620	Total Fumonisin (ppb)	0015	5,375	1,464	5,823	1,695	512.7	33	-0.26	620.99	0
620	Total Fumonisin (ppb)	0682	5,400	0.0000	5,823	1,695	512.7	33	-0.25	620.23	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
620	Total Fumonisin (ppb)	0202	5,489	257.0	5,823	1,695	512.7	33	-0.20	620.09	0
620	Total Fumonisin (ppb)	0227	5,500	1,000	5,823	1,695	512.7	33	-0.19	620.01	0
620	Total Fumonisin (ppb)	2226	5,525	470.0	5,823	1,695	512.7	33	-0.18	620.09	0
620	Total Fumonisin (ppb)	0027	5,605	550.0	5,823	1,695	512.7	33	-0.13	620.01	0
620	Total Fumonisin (ppb)	2113	5,650	100.0	5,823	1,695	512.7	33	-0.10	620.01	0
620	Total Fumonisin (ppb)	0870	5,675	570.0	5,823	1,695	512.7	33	-0.09	620.02	0
620	Total Fumonisin (ppb)	0511	5,700	600.0	5,823	1,695	512.7	33	-0.07	620.01	0
620	Total Fumonisin (ppb)	2033	5,955	38.00	5,823	1,695	512.7	33	0.08	620.23	0
620	Total Fumonisin (ppb)	2192	6,260	200.0	5,823	1,695	512.7	33	0.26	620.01	0
620	Total Fumonisin (ppb)	0208	6,300	1,000	5,823	1,695	512.7	33	0.28	620.20	0
620	Total Fumonisin (ppb)	0407	6,340	440.0	5,823	1,695	512.7	33	0.31	620.09	0
620	Total Fumonisin (ppb)	0353	6,402	122.0	5,823	1,695	512.7	33	0.34	620.23	0
620	Total Fumonisin (ppb)	0032	6,670	260.0	5,823	1,695	512.7	33	0.50	620.23	0
620	Total Fumonisin (ppb)	0066	6,890	220.0	5,823	1,695	512.7	33	0.63	620.09	0
620	Total Fumonisin (ppb)	0160	7,188	2,396	5,823	1,695	512.7	33	0.81	620.23	0
620	Total Fumonisin (ppb)	0142	7,550	700.0	5,823	1,695	512.7	33	1.02	620.98	0
620	Total Fumonisin (ppb)	0035	8,042	324.0	5,823	1,695	512.7	33	1.31	620.09	0
620	Total Fumonisin (ppb)	2269	8,121	1,898	5,823	1,695	512.7	33	1.36	620.98	0
620	Total Fumonisin (ppb)	2302	8,788	321.0	5,823	1,695	512.7	33	1.75	620.99	0
620	Total Fumonisin (ppb)	2022	11,310	340.0	5,823	1,695	512.7	33	3.24	620.07	0
621	Fumonsin B1 (ppb)	0098	2,600	200.0	4,195	1,232	316.6	19	-1.29	621.21	0
621	Fumonsin B1 (ppb)	2053	2,706	266.0	4,195	1,232	316.6	19	-1.21	621.23	0
621	Fumonsin B1 (ppb)	2207	3,300	400.0	4,195	1,232	316.6	19	-0.73	621.23	0
621	Fumonsin B1 (ppb)	0910	3,307	0.0000	4,195	1,232	316.6	19	-0.72	621.21	0
621	Fumonsin B1 (ppb)	0010	3,700	0.0000	4,195	1,232	316.6	19	-0.40	621.23	0
621	Fumonsin B1 (ppb)	0870	3,870	140.0	4,195	1,232	316.6	19	-0.26	621.23	0
621	Fumonsin B1 (ppb)	0619	3,876	219.0	4,195	1,232	316.6	19	-0.26	621.23	0
621	Fumonsin B1 (ppb)	2033	3,978	55.00	4,195	1,232	316.6	19	-0.18	621.23	0
621	Fumonsin B1 (ppb)	2136	4,135	230.0	4,195	1,232	316.6	19	-0.05	621.23	0
621	Fumonsin B1 (ppb)	0913	4,166	59.00	4,195	1,232	316.6	19	-0.02	621.23	0
621	Fumonsin B1 (ppb)	0682	4,200	0.0000	4,195	1,232	316.6	19	0.00	621.23	0
621	Fumonsin B1 (ppb)	0353	4,314	139.0	4,195	1,232	316.6	19	0.10	621.23	0
621	Fumonsin B1 (ppb)	0218	4,560	260.0	4,195	1,232	316.6	19	0.30	621.23	0
621	Fumonsin B1 (ppb)	0032	4,710	260.0	4,195	1,232	316.6	19	0.42	621.23	0
621	Fumonsin B1 (ppb)	2140	4,710	380.0	4,195	1,232	316.6	19	0.42	621.23	0
621	Fumonsin B1 (ppb)	0918	4,927	151.9	4,195	1,232	316.6	19	0.59	621.23	0
621	Fumonsin B1 (ppb)	0160	5,172	1,708	4,195	1,232	316.6	19	0.79	621.23	0
621	Fumonsin B1 (ppb)	0563	5,938	765.0	4,195	1,232	316.6	19	1.41	621.20	0
621	Fumonsin B1 (ppb)	2327	6,957	782.5	4,195	1,232	316.6	19	2.24	621.23	0
622	Fumonisin B2 (ppb)	2207	550.0	100.0	1,186	360.4	103.1	18	-1.76	622.23	0
622	Fumonisin B2 (ppb)	0098	650.0	100.0	1,186	360.4	103.1	18	-1.49	622.21	0
622	Fumonisin B2 (ppb)	2053	776.6	48.60	1,186	360.4	103.1	18	-1.14	622.23	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
622	Fumonisin B2 (ppb)	0910	879.0	0.0000	1,186	360.4	103.1	18	-0.85	622.21	0
622	Fumonisin B2 (ppb)	2136	970.0	100.0	1,186	360.4	103.1	18	-0.60	622.23	0
622	Fumonisin B2 (ppb)	0010	1,000	0.0000	1,186	360.4	103.1	18	-0.52	622.23	0
622	Fumonisin B2 (ppb)	0619	1,007	67.00	1,186	360.4	103.1	18	-0.50	622.23	0
622	Fumonisin B2 (ppb)	0870	1,150	20.00	1,186	360.4	103.1	18	-0.10	622.23	0
622	Fumonisin B2 (ppb)	0682	1,200	0.0000	1,186	360.4	103.1	18	0.04	622.23	0
622	Fumonisin B2 (ppb)	0218	1,250	60.00	1,186	360.4	103.1	18	0.18	622.23	0
622	Fumonisin B2 (ppb)	2140	1,288	120.0	1,186	360.4	103.1	18	0.28	622.23	0
622	Fumonisin B2 (ppb)	0913	1,303	26.00	1,186	360.4	103.1	18	0.33	622.23	0
622	Fumonisin B2 (ppb)	0353	1,363	103.0	1,186	360.4	103.1	18	0.49	622.23	0
622	Fumonisin B2 (ppb)	0032	1,395	10.00	1,186	360.4	103.1	18	0.58	622.23	0
622	Fumonisin B2 (ppb)	2033	1,403	5.000	1,186	360.4	103.1	18	0.60	622.23	0
622	Fumonisin B2 (ppb)	2327	1,637	414.2	1,186	360.4	103.1	18	1.25	622.23	0
622	Fumonisin B2 (ppb)	0160	1,704	539.0	1,186	360.4	103.1	18	1.44	622.23	0
622	Fumonisin B2 (ppb)	0918	2,360	142.7	1,186	360.4	103.1	18	3.26	622.23	0
623	Fumonisin B3 (ppb)	2053	282.4	18.60	483.6	150.6	37.12	9	-1.34	623.23	0
623	Fumonisin B3 (ppb)	0160	312.5	149.0	483.6	150.6	37.12	9	-1.14	623.23	0
623	Fumonisin B3 (ppb)	0870	450.0	20.00	483.6	150.6	37.12	9	-0.22	623.23	0
623	Fumonisin B3 (ppb)	0010	500.0	0.0000	483.6	150.6	37.12	9	0.11	623.23	0
623	Fumonisin B3 (ppb)	2327	510.2	9.496	483.6	150.6	37.12	9	0.18	623.23	0
623	Fumonisin B3 (ppb)	0913	534.0	20.00	483.6	150.6	37.12	9	0.33	623.23	0
623	Fumonisin B3 (ppb)	0032	566.0	4.000	483.6	150.6	37.12	9	0.55	623.23	0
623	Fumonisin B3 (ppb)	2033	575.0	12.00	483.6	150.6	37.12	9	0.61	623.23	0
623	Fumonisin B3 (ppb)	0353	615.5	101.0	483.6	150.6	37.12	9	0.88	623.23	0
630	Ochratoxin A (ppb)	0227	21.00	2.000	100.6	32.68	11.75	34	-2.43	630.01	0
630	Ochratoxin A (ppb)	2327	28.30	4.458	100.6	32.68	11.75	34	-2.21	630.23	0
630	Ochratoxin A (ppb)	0407	37.03	3.190	100.6	32.68	11.75	34	-1.94	630.04	0
630	Ochratoxin A (ppb)	2269	58.45	24.10	100.6	32.68	11.75	34	-1.29	630.98	0
630	Ochratoxin A (ppb)	0563	58.50	3.000	100.6	32.68	11.75	34	-1.29	630.20	0
630	Ochratoxin A (ppb)	2098	71.80	5.800	100.6	32.68	11.75	34	-0.88	630.01	0
630	Ochratoxin A (ppb)	0015	73.50	21.00	100.6	32.68	11.75	34	-0.83	630.99	0
630	Ochratoxin A (ppb)	2226	73.83	10.64	100.6	32.68	11.75	34	-0.82	630.05	0
630	Ochratoxin A (ppb)	0870	76.40	2.000	100.6	32.68	11.75	34	-0.74	630.23	0
630	Ochratoxin A (ppb)	0910	77.21	46.01	100.6	32.68	11.75	34	-0.71	630.21	0
630	Ochratoxin A (ppb)	0066	80.65	1.100	100.6	32.68	11.75	34	-0.61	630.05	0
630	Ochratoxin A (ppb)	2192	85.96	18.65	100.6	32.68	11.75	34	-0.45	630.20	0
630	Ochratoxin A (ppb)	0682	86.00	0.0000	100.6	32.68	11.75	34	-0.45	630.23	0
630	Ochratoxin A (ppb)	2196	88.74	0.0000	100.6	32.68	11.75	34	-0.36	630.20	0
630	Ochratoxin A (ppb)	2053	89.35	2.660	100.6	32.68	11.75	34	-0.34	630.23	0
630	Ochratoxin A (ppb)	0913	91.55	1.300	100.6	32.68	11.75	34	-0.28	630.23	0
630	Ochratoxin A (ppb)	2302	93.54	4.258	100.6	32.68	11.75	34	-0.21	630.99	0
630	Ochratoxin A (ppb)	0042	94.58	29.59	100.6	32.68	11.75	34	-0.18	630.98	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
630	Ochratoxin A (ppb)	2033	99.50	21.00	100.6	32.68	11.75	34	-0.03	630.23	0
630	Ochratoxin A (ppb)	2022	99.77	1.980	100.6	32.68	11.75	34	-0.02	630.03	0
630	Ochratoxin A (ppb)	0619	110.5	9.000	100.6	32.68	11.75	34	0.30	630.23	0
630	Ochratoxin A (ppb)	0032	123.5	7.000	100.6	32.68	11.75	34	0.70	630.23	0
630	Ochratoxin A (ppb)	0218	125.0	0.0000	100.6	32.68	11.75	34	0.75	630.23	0
630	Ochratoxin A (ppb)	0353	125.5	3.000	100.6	32.68	11.75	34	0.76	630.23	0
630	Ochratoxin A (ppb)	0918	126.9	3.100	100.6	32.68	11.75	34	0.80	630.23	0
630	Ochratoxin A (ppb)	2207	131.5	15.00	100.6	32.68	11.75	34	0.95	630.23	0
630	Ochratoxin A (ppb)	0051	132.0	4.000	100.6	32.68	11.75	34	0.96	630.20	0
630	Ochratoxin A (ppb)	2140	134.0	12.00	100.6	32.68	11.75	34	1.02	630.23	0
630	Ochratoxin A (ppb)	0870	139.8	18.50	100.6	32.68	11.75	34	1.20	630.98	0
630	Ochratoxin A (ppb)	0160	140.0	60.00	100.6	32.68	11.75	34	1.21	630.23	0
630	Ochratoxin A (ppb)	2136	141.5	27.00	100.6	32.68	11.75	34	1.25	630.23	0
630	Ochratoxin A (ppb)	0098	146.5	11.00	100.6	32.68	11.75	34	1.41	630.20	0
630	Ochratoxin A (ppb)	0010	153.5	5.000	100.6	32.68	11.75	34	1.62	630.23	0
630	Ochratoxin A (ppb)	0964	174.9	22.20	100.6	32.68	11.75	34	2.28	630.23	0
630	Ochratoxin A (ppb)	0529	507.0	6.000	100.6	32.68	11.75	34	12.44	630.01	2
640	T-2 Toxin (ppb)	0563	73.50	9.000	146.1	47.01	23.66	23	-1.54	640.25	0
640	T-2 Toxin (ppb)	0910	79.20	17.60	146.1	47.01	23.66	23	-1.42	640.21	0
640	T-2 Toxin (ppb)	2033	100.0	0.0000	146.1	47.01	23.66	23	-0.98	640.23	0
640	T-2 Toxin (ppb)	0682	107.0	0.0000	146.1	47.01	23.66	23	-0.83	640.23	0
640	T-2 Toxin (ppb)	0619	114.0	0.0000	146.1	47.01	23.66	23	-0.68	640.23	0
640	T-2 Toxin (ppb)	0913	116.5	15.40	146.1	47.01	23.66	23	-0.63	640.23	0
640	T-2 Toxin (ppb)	2053	123.6	8.700	146.1	47.01	23.66	23	-0.48	640.23	0
640	T-2 Toxin (ppb)	2327	128.3	18.00	146.1	47.01	23.66	23	-0.38	640.23	0
640	T-2 Toxin (ppb)	2136	131.0	22.00	146.1	47.01	23.66	23	-0.32	640.23	0
640	T-2 Toxin (ppb)	0407	133.0	3.826	146.1	47.01	23.66	23	-0.28	640.05	0
640	T-2 Toxin (ppb)	0918	133.6	10.41	146.1	47.01	23.66	23	-0.27	640.23	0
640	T-2 Toxin (ppb)	0353	134.0	2.000	146.1	47.01	23.66	23	-0.26	640.23	0
640	T-2 Toxin (ppb)	0870	142.4	14.58	146.1	47.01	23.66	23	-0.08	640.23	0
640	T-2 Toxin (ppb)	2140	154.0	10.00	146.1	47.01	23.66	23	0.17	640.23	0
640	T-2 Toxin (ppb)	2129	156.5	33.00	146.1	47.01	23.66	23	0.22	640.23	0
640	T-2 Toxin (ppb)	0160	160.5	33.00	146.1	47.01	23.66	23	0.31	640.23	0
640	T-2 Toxin (ppb)	2207	165.0	50.00	146.1	47.01	23.66	23	0.40	640.23	0
640	T-2 Toxin (ppb)	0015	171.0	40.00	146.1	47.01	23.66	23	0.53	640.99	0
640	T-2 Toxin (ppb)	0218	192.0	4.000	146.1	47.01	23.66	23	0.98	640.23	0
640	T-2 Toxin (ppb)	0066	195.4	2.670	146.1	47.01	23.66	23	1.05	640.03	0
640	T-2 Toxin (ppb)	2302	224.5	25.00	146.1	47.01	23.66	23	1.67	640.99	0
640	T-2 Toxin (ppb)	0042	226.4	154.1	146.1	47.01	23.66	23	1.71	640.98	0
640	T-2 Toxin (ppb)	2269	242.8	71.00	146.1	47.01	23.66	23	2.06	640.98	0
641	HT-2 Toxin (ppb)	0910	41.75	9.900	94.61	30.80	5.458	12	-1.72	641.21	0
641	HT-2 Toxin (ppb)	2327	65.26	3.786	94.61	30.80	5.458	12	-0.95	641.23	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
641	HT-2 Toxin (ppb)	0918	68.74	0.4800	94.61	30.80	5.458	12	-0.84	641.23	0
641	HT-2 Toxin (ppb)	0870	81.37	5.020	94.61	30.80	5.458	12	-0.43	641.23	0
641	HT-2 Toxin (ppb)	0353	82.00	4.000	94.61	30.80	5.458	12	-0.41	641.23	0
641	HT-2 Toxin (ppb)	2033	85.90	6.800	94.61	30.80	5.458	12	-0.28	641.23	0
641	HT-2 Toxin (ppb)	0913	90.25	4.700	94.61	30.80	5.458	12	-0.14	641.23	0
641	HT-2 Toxin (ppb)	2136	103.0	26.00	94.61	30.80	5.458	12	0.27	641.23	0
641	HT-2 Toxin (ppb)	0218	118.0	0.0000	94.61	30.80	5.458	12	0.76	641.23	0
641	HT-2 Toxin (ppb)	2053	121.6	1.500	94.61	30.80	5.458	12	0.87	641.23	0
641	HT-2 Toxin (ppb)	2207	130.0	0.0000	94.61	30.80	5.458	12	1.15	641.23	0
641	HT-2 Toxin (ppb)	0964	178.8	3.310	94.61	30.80	5.458	12	2.73	641.23	0
641	HT-2 Toxin (ppb)	0619	0.0000	0.0000	94.61	30.80	5.458	12		641.23	4
641	HT-2 Toxin (ppb)	2140	< 200		94.61	30.80	5.458	12		641.23	5
642	T-2 Toxin + HT-2 Toxin (ppb)	0407	163.8	3.240	303.2	95.61	24.58	11	-1.46	642.99	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0227	190.5	5.000	303.2	95.61	24.58	11	-1.18	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0142	202.3	32.60	303.2	95.61	24.58	11	-1.05	642.98	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0353	226.0	14.00	303.2	95.61	24.58	11	-0.81	642.04	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0590	249.0	56.00	303.2	95.61	24.58	11	-0.57	642.98	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2302	321.1	0.8560	303.2	95.61	24.58	11	0.19	642.99	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2269	321.6	71.00	303.2	95.61	24.58	11	0.19	642.98	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0870	332.0	24.90	303.2	95.61	24.58	11	0.30	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2098	438.9	11.90	303.2	95.61	24.58	11	1.42	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2022	443.7	8.470	303.2	95.61	24.58	11	1.47	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0529	445.8	42.40	303.2	95.61	24.58	11	1.49	642.21	0
650	Zearalenone (ppb)	2169	32.95	0.7000	300.1	94.66	37.77	42	-2.82	650.98	0
650	Zearalenone (ppb)	0520	90.50	3.000	300.1	94.66	37.77	42	-2.21	650.21	0
650	Zearalenone (ppb)	0202	163.9	30.90	300.1	94.66	37.77	42	-1.44	650.06	0
650	Zearalenone (ppb)	2192	179.0	56.49	300.1	94.66	37.77	42	-1.28	650.20	0
650	Zearalenone (ppb)	0032	179.5	5.000	300.1	94.66	37.77	42	-1.27	650.24	0
650	Zearalenone (ppb)	2098	195.3	17.00	300.1	94.66	37.77	42	-1.11	650.01	0
650	Zearalenone (ppb)	2147	210.0	0.0000	300.1	94.66	37.77	42	-0.95	650.98	0
650	Zearalenone (ppb)	0870	221.3	2.832	300.1	94.66	37.77	42	-0.83	650.24	0
650	Zearalenone (ppb)	2196	221.7	0.0000	300.1	94.66	37.77	42	-0.83	650.20	0
650	Zearalenone (ppb)	0098	225.0	30.00	300.1	94.66	37.77	42	-0.79	650.20	0
650	Zearalenone (ppb)	2053	229.5	8.700	300.1	94.66	37.77	42	-0.75	650.24	0
650	Zearalenone (ppb)	0066	235.4	10.07	300.1	94.66	37.77	42	-0.68	650.07	0
650	Zearalenone (ppb)	2226	238.5	20.46	300.1	94.66	37.77	42	-0.65	650.06	0
650	Zearalenone (ppb)	2207	240.0	60.00	300.1	94.66	37.77	42	-0.63	650.24	0
650	Zearalenone (ppb)	0913	261.8	19.10	300.1	94.66	37.77	42	-0.40	650.24	0
650	Zearalenone (ppb)	2140	271.5	93.00	300.1	94.66	37.77	42	-0.30	650.24	0
650	Zearalenone (ppb)	0042	278.5	22.34	300.1	94.66	37.77	42	-0.23	650.98	0
650	Zearalenone (ppb)	0353	283.0	54.00	300.1	94.66	37.77	42	-0.18	650.23	0
650	Zearalenone (ppb)	0051	289.0	48.00	300.1	94.66	37.77	42	-0.12	650.20	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
650	Zearalenone (ppb)	0010	290.0	80.00	300.1	94.66	37.77	42	-0.11	650.24	0
650	Zearalenone (ppb)	2066	291.5	33.00	300.1	94.66	37.77	42	-0.09	650.06	0
650	Zearalenone (ppb)	0918	292.8	25.30	300.1	94.66	37.77	42	-0.08	650.24	0
650	Zearalenone (ppb)	0619	304.0	36.00	300.1	94.66	37.77	42	0.04	650.24	0
650	Zearalenone (ppb)	0015	308.5	109.0	300.1	94.66	37.77	42	0.09	650.99	0
650	Zearalenone (ppb)	2129	321.5	5.000	300.1	94.66	37.77	42	0.23	650.24	0
650	Zearalenone (ppb)	2136	324.5	35.00	300.1	94.66	37.77	42	0.26	650.24	0
650	Zearalenone (ppb)	0218	327.0	10.00	300.1	94.66	37.77	42	0.28	650.24	0
650	Zearalenone (ppb)	2269	336.0	30.00	300.1	94.66	37.77	42	0.38	650.98	0
650	Zearalenone (ppb)	2113	340.0	120.0	300.1	94.66	37.77	42	0.42	650.01	0
650	Zearalenone (ppb)	0910	347.2	36.10	300.1	94.66	37.77	42	0.50	650.20	0
650	Zearalenone (ppb)	0563	356.5	9.000	300.1	94.66	37.77	42	0.60	650.20	0
650	Zearalenone (ppb)	0035	365.0	63.65	300.1	94.66	37.77	42	0.69	650.07	0
650	Zearalenone (ppb)	0529	381.1	3.200	300.1	94.66	37.77	42	0.86	650.01	0
650	Zearalenone (ppb)	0407	381.4	35.44	300.1	94.66	37.77	42	0.86	650.07	0
650	Zearalenone (ppb)	2022	396.0	20.00	300.1	94.66	37.77	42	1.01	650.04	0
650	Zearalenone (ppb)	0142	397.1	20.20	300.1	94.66	37.77	42	1.03	650.98	0
650	Zearalenone (ppb)	0413	405.0	10.00	300.1	94.66	37.77	42	1.11	650.20	0
650	Zearalenone (ppb)	0870	429.9	45.40	300.1	94.66	37.77	42	1.37	650.01	0
650	Zearalenone (ppb)	2302	437.6	4.815	300.1	94.66	37.77	42	1.45	650.99	0
650	Zearalenone (ppb)	0027	448.3	120.0	300.1	94.66	37.77	42	1.57	650.01	0
650	Zearalenone (ppb)	0160	511.5	165.0	300.1	94.66	37.77	42	2.23	650.23	0
650	Zearalenone (ppb)	2327	583.1	88.75	300.1	94.66	37.77	42	2.99	650.24	0
650	Zearalenone (ppb)	0297	5,871	418.0	300.1	94.66	37.77	42	58.85	650.01	2

Notes: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2 and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = Rejected for duplicates too far apart, 2 = Rejected as outlier, 5 = A reporting limit and 4 = zeros submitted as values. Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). Flag 3 indicates not used in statistics.