



AAFCO
Association of American Feed Control Officials



Animal Feed Scheme
Dairy Calf Starter, Medicated
Test Material Code # 202231

Labs Reporting: 171
Analytes Reported 110
Issue Date : 12/31/2022

Analyte Proficiency Testing Report

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
000	Urea (%)	0875	2.230	0.0800	5.166	3.118	0.0657	7	-0.94	000.99	0
000	Urea (%)	0619	2.265	0.0100	5.166	3.118	0.0657	7	-0.93	000.99	0
000	Urea (%)	0019	2.320	0.0800	5.166	3.118	0.0657	7	-0.91	000.02	0
000	Urea (%)	0278	6.650	0.1000	5.166	3.118	0.0657	7	0.48	000.02	0
000	Urea (%)	0171	6.880	0.0000	5.166	3.118	0.0657	7	0.55	000.02	0
000	Urea (%)	0036	7.755	0.1896	5.166	3.118	0.0657	7	0.83	000.02	0
000	Urea (%)	2192	8.060	0.0000	5.166	3.118	0.0657	7	0.93	000.02	0
001	Loss on Drying (%)	0618	7.525	0.2500	9.478	0.3486	0.1353	76	-5.60	001.99	0
001	Loss on Drying (%)	2155	7.560	0.4000	9.478	0.3486	0.1353	76	-5.50	001.99	0
001	Loss on Drying (%)	0278	8.440	0.1600	9.478	0.3486	0.1353	76	-2.98	001.07	0
001	Loss on Drying (%)	2081	8.555	0.1500	9.478	0.3486	0.1353	76	-2.65	001.99	0
001	Loss on Drying (%)	0186	8.576	0.0020	9.478	0.3486	0.1353	76	-2.59	001.99	0
001	Loss on Drying (%)	0676	8.745	0.0700	9.478	0.3486	0.1353	76	-2.10	001.99	0
001	Loss on Drying (%)	0089	9.015	0.0300	9.478	0.3486	0.1353	76	-1.33	001.07	0
001	Loss on Drying (%)	0038	9.035	0.6900	9.478	0.3486	0.1353	76	-1.27	001.07	0
001	Loss on Drying (%)	0353	9.050	0.0600	9.478	0.3486	0.1353	76	-1.23	001.07	0
001	Loss on Drying (%)	0914	9.060	0.0600	9.478	0.3486	0.1353	76	-1.20	001.00	0
001	Loss on Drying (%)	0596	9.075	0.0500	9.478	0.3486	0.1353	76	-1.16	001.07	0
001	Loss on Drying (%)	2259	9.105	0.0080	9.478	0.3486	0.1353	76	-1.07	001.07	0
001	Loss on Drying (%)	2144	9.125	0.0100	9.478	0.3486	0.1353	76	-1.01	001.07	0
001	Loss on Drying (%)	2246	9.130	0.1000	9.478	0.3486	0.1353	76	-1.00	001.99	0
001	Loss on Drying (%)	0918	9.215	0.0500	9.478	0.3486	0.1353	76	-0.75	001.99	0
001	Loss on Drying (%)	0656	9.220	0.0000	9.478	0.3486	0.1353	76	-0.74	001.99	0
001	Loss on Drying (%)	0345	9.225	0.0100	9.478	0.3486	0.1353	76	-0.73	001.07	0
001	Loss on Drying (%)	0610	9.235	0.0300	9.478	0.3486	0.1353	76	-0.70	001.99	0
001	Loss on Drying (%)	0910	9.250	0.1000	9.478	0.3486	0.1353	76	-0.65	001.99	0
001	Loss on Drying (%)	2190	9.285	0.3700	9.478	0.3486	0.1353	76	-0.55	001.07	0
001	Loss on Drying (%)	0512	9.299	0.5560	9.478	0.3486	0.1353	76	-0.51	001.07	0
001	Loss on Drying (%)	0010	9.300	0.0000	9.478	0.3486	0.1353	76	-0.51	001.07	0
001	Loss on Drying (%)	0035	9.300	0.0600	9.478	0.3486	0.1353	76	-0.51	001.07	0
001	Loss on Drying (%)	0510	9.300	0.0000	9.478	0.3486	0.1353	76	-0.51	001.99	0
001	Loss on Drying (%)	0027	9.319	0.0270	9.478	0.3486	0.1353	76	-0.46	001.00	0
001	Loss on Drying (%)	0148	9.360	0.0800	9.478	0.3486	0.1353	76	-0.34	001.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
001	Loss on Drying (%)	0226	9.365	0.0900	9.478	0.3486	0.1353	76	-0.32	001.07	0
001	Loss on Drying (%)	0297	9.365	0.2300	9.478	0.3486	0.1353	76	-0.32	001.07	0
001	Loss on Drying (%)	0876	9.370	0.3629	9.478	0.3486	0.1353	76	-0.31	001.07	0
001	Loss on Drying (%)	0015	9.395	0.0100	9.478	0.3486	0.1353	76	-0.24	001.07	0
001	Loss on Drying (%)	0644	9.397	0.0310	9.478	0.3486	0.1353	76	-0.23	001.05	0
001	Loss on Drying (%)	0407	9.405	0.0929	9.478	0.3486	0.1353	76	-0.21	001.00	0
001	Loss on Drying (%)	0948	9.440	0.0400	9.478	0.3486	0.1353	76	-0.11	001.99	0
001	Loss on Drying (%)	0643	9.464	0.0140	9.478	0.3486	0.1353	76	-0.04	001.07	0
001	Loss on Drying (%)	0034	9.465	0.0700	9.478	0.3486	0.1353	76	-0.04	001.07	0
001	Loss on Drying (%)	0912	9.465	0.0700	9.478	0.3486	0.1353	76	-0.04	001.99	0
001	Loss on Drying (%)	0171	9.470	0.0400	9.478	0.3486	0.1353	76	-0.02	001.07	0
001	Loss on Drying (%)	2192	9.480	0.0400	9.478	0.3486	0.1353	76	0.01	001.07	0
001	Loss on Drying (%)	0650	9.500	0.1800	9.478	0.3486	0.1353	76	0.06	001.07	0
001	Loss on Drying (%)	0683	9.500	0.0200	9.478	0.3486	0.1353	76	0.06	001.07	0
001	Loss on Drying (%)	0066	9.505	0.1100	9.478	0.3486	0.1353	76	0.08	001.07	0
001	Loss on Drying (%)	0309	9.517	0.0960	9.478	0.3486	0.1353	76	0.11	001.00	0
001	Loss on Drying (%)	0571	9.530	0.0000	9.478	0.3486	0.1353	76	0.15	001.07	0
001	Loss on Drying (%)	0049	9.540	0.2000	9.478	0.3486	0.1353	76	0.18	001.07	0
001	Loss on Drying (%)	0366	9.550	0.3000	9.478	0.3486	0.1353	76	0.21	001.07	0
001	Loss on Drying (%)	0505	9.550	0.3800	9.478	0.3486	0.1353	76	0.21	001.07	0
001	Loss on Drying (%)	2059	9.550	0.0200	9.478	0.3486	0.1353	76	0.21	001.07	0
001	Loss on Drying (%)	0003	9.555	0.1100	9.478	0.3486	0.1353	76	0.22	001.07	0
001	Loss on Drying (%)	2009	9.557	0.0200	9.478	0.3486	0.1353	76	0.23	001.07	0
001	Loss on Drying (%)	2073	9.560	0.0200	9.478	0.3486	0.1353	76	0.23	001.00	0
001	Loss on Drying (%)	0675	9.595	0.2700	9.478	0.3486	0.1353	76	0.34	001.07	0
001	Loss on Drying (%)	0045	9.600	0.6000	9.478	0.3486	0.1353	76	0.35	001.07	0
001	Loss on Drying (%)	0689	9.650	0.1000	9.478	0.3486	0.1353	76	0.49	001.07	0
001	Loss on Drying (%)	2319	9.650	0.1000	9.478	0.3486	0.1353	76	0.49	001.99	0
001	Loss on Drying (%)	0083	9.655	0.0100	9.478	0.3486	0.1353	76	0.51	001.07	0
001	Loss on Drying (%)	0098	9.660	0.0200	9.478	0.3486	0.1353	76	0.52	001.07	0
001	Loss on Drying (%)	0169	9.675	0.0300	9.478	0.3486	0.1353	76	0.56	001.00	0
001	Loss on Drying (%)	2109	9.690	0.0400	9.478	0.3486	0.1353	76	0.61	001.07	0
001	Loss on Drying (%)	2012	9.710	0.0000	9.478	0.3486	0.1353	76	0.67	001.07	0
001	Loss on Drying (%)	2268	9.710	0.2600	9.478	0.3486	0.1353	76	0.67	001.99	0
001	Loss on Drying (%)	0028	9.750	0.1000	9.478	0.3486	0.1353	76	0.78	001.07	0
001	Loss on Drying (%)	0142	9.780	0.0600	9.478	0.3486	0.1353	76	0.87	001.07	0
001	Loss on Drying (%)	0619	9.815	0.1100	9.478	0.3486	0.1353	76	0.97	001.03	0
001	Loss on Drying (%)	0885	9.824	0.0015	9.478	0.3486	0.1353	76	0.99	001.07	0
001	Loss on Drying (%)	0019	9.825	0.0700	9.478	0.3486	0.1353	76	1.00	001.99	0
001	Loss on Drying (%)	2325	9.830	0.0200	9.478	0.3486	0.1353	76	1.01	001.99	0
001	Loss on Drying (%)	0581	9.885	0.0300	9.478	0.3486	0.1353	76	1.17	001.07	0
001	Loss on Drying (%)	0504	9.895	0.9300	9.478	0.3486	0.1353	76	1.20	001.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
001	Loss on Drying (%)	0629	10.10	0.0300	9.478	0.3486	0.1353	76	1.77	001.99	0
001	Loss on Drying (%)	0407	10.15	0.1614	9.478	0.3486	0.1353	76	1.94	001.07	0
001	Loss on Drying (%)	0074	10.18	0.5500	9.478	0.3486	0.1353	76	2.00	001.07	0
001	Loss on Drying (%)	0693	10.22	0.3200	9.478	0.3486	0.1353	76	2.13	001.07	0
001	Loss on Drying (%)	0357	10.48	0.1166	9.478	0.3486	0.1353	76	2.87	001.99	0
001	Loss on Drying (%)	0682	10.80	0.0000	9.478	0.3486	0.1353	76	3.79	001.07	0
001	Loss on Drying (%)	0843	11.16	0.2400	9.478	0.3486	0.1353	76	4.83	001.03	0
001	Loss on Drying (%)	0843	11.16	0.2400	9.478	0.3486	0.1353	76	4.83	001.07	0
001	Loss on Drying (%)	0638	7.200	4.400	9.478	0.3486	0.1353	76	-6.54	001.07	1
001	Loss on Drying (%)	0560	0.0931	0.0008	9.478	0.3486	0.1353	76	-26.92	001.99	2
002	Protein, Crude (%)	0859	18.04	0.3670	24.09	0.3490	0.1733	168	-17.36	002.06	0
002	Protein, Crude (%)	0297	20.33	0.1100	24.09	0.3490	0.1733	168	-10.80	002.11	0
002	Protein, Crude (%)	0553	20.74	0.2000	24.09	0.3490	0.1733	168	-9.61	002.11	0
002	Protein, Crude (%)	0885	21.21	0.0200	24.09	0.3490	0.1733	168	-8.27	002.11	0
002	Protein, Crude (%)	0169	23.11	0.0700	24.09	0.3490	0.1733	168	-2.84	002.00	0
002	Protein, Crude (%)	2073	23.11	0.1100	24.09	0.3490	0.1733	168	-2.84	002.05	0
002	Protein, Crude (%)	2199	23.15	0.1000	24.09	0.3490	0.1733	168	-2.71	002.05	0
002	Protein, Crude (%)	0596	23.30	0.0700	24.09	0.3490	0.1733	168	-2.29	002.05	0
002	Protein, Crude (%)	0345	23.45	0.1000	24.09	0.3490	0.1733	168	-1.85	002.06	0
002	Protein, Crude (%)	2144	23.48	0.1600	24.09	0.3490	0.1733	168	-1.76	002.05	0
002	Protein, Crude (%)	0042	23.50	0.0000	24.09	0.3490	0.1733	168	-1.70	002.06	0
002	Protein, Crude (%)	2006	23.51	0.1300	24.09	0.3490	0.1733	168	-1.69	002.05	0
002	Protein, Crude (%)	2200	23.55	0.3000	24.09	0.3490	0.1733	168	-1.56	002.05	0
002	Protein, Crude (%)	2190	23.56	0.3900	24.09	0.3490	0.1733	168	-1.55	002.05	0
002	Protein, Crude (%)	0651	23.56	0.1129	24.09	0.3490	0.1733	168	-1.54	002.05	0
002	Protein, Crude (%)	2188	23.59	0.0000	24.09	0.3490	0.1733	168	-1.45	002.01	0
002	Protein, Crude (%)	0643	23.60	0.0300	24.09	0.3490	0.1733	168	-1.43	002.05	0
002	Protein, Crude (%)	0202	23.67	0.3200	24.09	0.3490	0.1733	168	-1.22	002.06	0
002	Protein, Crude (%)	2259	23.68	0.4400	24.09	0.3490	0.1733	168	-1.19	002.01	0
002	Protein, Crude (%)	0692	23.70	0.4000	24.09	0.3490	0.1733	168	-1.13	002.06	0
002	Protein, Crude (%)	0598	23.71	0.1500	24.09	0.3490	0.1733	168	-1.12	002.06	0
002	Protein, Crude (%)	2325	23.73	0.0200	24.09	0.3490	0.1733	168	-1.05	002.01	0
002	Protein, Crude (%)	2246	23.74	0.2600	24.09	0.3490	0.1733	168	-1.02	002.05	0
002	Protein, Crude (%)	0723	23.76	0.0600	24.09	0.3490	0.1733	168	-0.96	002.01	0
002	Protein, Crude (%)	0164	23.79	0.3100	24.09	0.3490	0.1733	168	-0.89	002.01	0
002	Protein, Crude (%)	0885	23.79	0.0324	24.09	0.3490	0.1733	168	-0.88	002.05	0
002	Protein, Crude (%)	2193	23.80	0.1100	24.09	0.3490	0.1733	168	-0.86	002.01	0
002	Protein, Crude (%)	2022	23.80	0.0700	24.09	0.3490	0.1733	168	-0.86	002.06	0
002	Protein, Crude (%)	0910	23.80	0.4000	24.09	0.3490	0.1733	168	-0.85	002.01	0
002	Protein, Crude (%)	2022	23.81	0.0200	24.09	0.3490	0.1733	168	-0.82	002.05	0
002	Protein, Crude (%)	0354	23.82	0.0500	24.09	0.3490	0.1733	168	-0.80	002.05	0
002	Protein, Crude (%)	0028	23.82	0.0400	24.09	0.3490	0.1733	168	-0.79	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	2059	23.82	0.1875	24.09	0.3490	0.1733	168	-0.77	002.06	0
002	Protein, Crude (%)	0036	23.83	0.0750	24.09	0.3490	0.1733	168	-0.77	002.06	0
002	Protein, Crude (%)	0726	23.83	0.0800	24.09	0.3490	0.1733	168	-0.76	002.06	0
002	Protein, Crude (%)	0675	23.84	0.0900	24.09	0.3490	0.1733	168	-0.74	002.06	0
002	Protein, Crude (%)	0297	23.85	0.0900	24.09	0.3490	0.1733	168	-0.72	002.06	0
002	Protein, Crude (%)	0689	23.85	0.3000	24.09	0.3490	0.1733	168	-0.70	002.05	0
002	Protein, Crude (%)	2349	23.85	0.3000	24.09	0.3490	0.1733	168	-0.70	002.06	0
002	Protein, Crude (%)	2192	23.86	0.0600	24.09	0.3490	0.1733	168	-0.67	002.05	0
002	Protein, Crude (%)	2109	23.87	0.0700	24.09	0.3490	0.1733	168	-0.66	002.05	0
002	Protein, Crude (%)	0004	23.87	0.0700	24.09	0.3490	0.1733	168	-0.66	002.06	0
002	Protein, Crude (%)	0263	23.87	0.0040	24.09	0.3490	0.1733	168	-0.64	002.06	0
002	Protein, Crude (%)	0848	23.88	0.2500	24.09	0.3490	0.1733	168	-0.63	002.06	0
002	Protein, Crude (%)	0848	23.88	0.0400	24.09	0.3490	0.1733	168	-0.62	002.01	0
002	Protein, Crude (%)	2155	23.89	0.6480	24.09	0.3490	0.1733	168	-0.60	002.06	0
002	Protein, Crude (%)	0278	23.90	0.0000	24.09	0.3490	0.1733	168	-0.56	002.06	0
002	Protein, Crude (%)	0870	23.90	0.0250	24.09	0.3490	0.1733	168	-0.56	002.06	0
002	Protein, Crude (%)	0910	23.90	0.0000	24.09	0.3490	0.1733	168	-0.56	002.06	0
002	Protein, Crude (%)	2150	23.90	0.1000	24.09	0.3490	0.1733	168	-0.56	002.06	0
002	Protein, Crude (%)	0539	23.91	0.2300	24.09	0.3490	0.1733	168	-0.54	002.06	0
002	Protein, Crude (%)	2397	23.92	0.0900	24.09	0.3490	0.1733	168	-0.52	002.01	0
002	Protein, Crude (%)	0019	23.92	0.2200	24.09	0.3490	0.1733	168	-0.50	002.06	0
002	Protein, Crude (%)	0529	23.93	0.0300	24.09	0.3490	0.1733	168	-0.49	002.06	0
002	Protein, Crude (%)	0619	23.95	0.1000	24.09	0.3490	0.1733	168	-0.42	002.05	0
002	Protein, Crude (%)	0010	23.95	0.1000	24.09	0.3490	0.1733	168	-0.42	002.06	0
002	Protein, Crude (%)	0294	23.95	0.1000	24.09	0.3490	0.1733	168	-0.42	002.06	0
002	Protein, Crude (%)	0610	23.95	0.1000	24.09	0.3490	0.1733	168	-0.42	002.06	0
002	Protein, Crude (%)	0619	23.95	0.1000	24.09	0.3490	0.1733	168	-0.42	002.06	0
002	Protein, Crude (%)	0652	23.95	0.3000	24.09	0.3490	0.1733	168	-0.42	002.01	0
002	Protein, Crude (%)	0504	23.96	0.1900	24.09	0.3490	0.1733	168	-0.40	002.04	0
002	Protein, Crude (%)	0870	23.97	0.3410	24.09	0.3490	0.1733	168	-0.37	002.01	0
002	Protein, Crude (%)	0144	23.97	0.0400	24.09	0.3490	0.1733	168	-0.36	002.06	0
002	Protein, Crude (%)	0504	23.97	0.1200	24.09	0.3490	0.1733	168	-0.36	002.06	0
002	Protein, Crude (%)	2146	23.98	0.2800	24.09	0.3490	0.1733	168	-0.33	002.05	0
002	Protein, Crude (%)	0505	23.99	0.0700	24.09	0.3490	0.1733	168	-0.31	002.06	0
002	Protein, Crude (%)	2340	24.00	0.0010	24.09	0.3490	0.1733	168	-0.28	002.05	0
002	Protein, Crude (%)	2319	24.00	0.0000	24.09	0.3490	0.1733	168	-0.27	002.06	0
002	Protein, Crude (%)	2174	24.02	0.4800	24.09	0.3490	0.1733	168	-0.21	002.06	0
002	Protein, Crude (%)	0003	24.03	0.2100	24.09	0.3490	0.1733	168	-0.20	002.06	0
002	Protein, Crude (%)	2337	24.03	0.0000	24.09	0.3490	0.1733	168	-0.19	002.05	0
002	Protein, Crude (%)	0100	24.03	0.4000	24.09	0.3490	0.1733	168	-0.19	002.06	0
002	Protein, Crude (%)	0676	24.03	0.0600	24.09	0.3490	0.1733	168	-0.19	002.06	0
002	Protein, Crude (%)	0953	24.04	0.0900	24.09	0.3490	0.1733	168	-0.17	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0939	24.06	0.2100	24.09	0.3490	0.1733	168	-0.11	002.05	0
002	Protein, Crude (%)	2168	24.06	0.2000	24.09	0.3490	0.1733	168	-0.10	002.06	0
002	Protein, Crude (%)	2196	24.07	0.0000	24.09	0.3490	0.1733	168	-0.07	002.01	0
002	Protein, Crude (%)	0108	24.08	0.0100	24.09	0.3490	0.1733	168	-0.06	002.06	0
002	Protein, Crude (%)	0066	24.08	0.1300	24.09	0.3490	0.1733	168	-0.06	002.02	0
002	Protein, Crude (%)	2302	24.09	0.0000	24.09	0.3490	0.1733	168	-0.01	002.11	0
002	Protein, Crude (%)	0644	24.10	0.0450	24.09	0.3490	0.1733	168	0.00	002.06	0
002	Protein, Crude (%)	0047	24.10	0.4000	24.09	0.3490	0.1733	168	0.01	002.06	0
002	Protein, Crude (%)	0049	24.10	0.4000	24.09	0.3490	0.1733	168	0.01	002.06	0
002	Protein, Crude (%)	0510	24.10	1.000	24.09	0.3490	0.1733	168	0.01	002.06	0
002	Protein, Crude (%)	0148	24.11	0.9140	24.09	0.3490	0.1733	168	0.03	002.06	0
002	Protein, Crude (%)	0629	24.12	0.0200	24.09	0.3490	0.1733	168	0.07	002.10	0
002	Protein, Crude (%)	0948	24.12	0.0000	24.09	0.3490	0.1733	168	0.07	002.06	0
002	Protein, Crude (%)	0520	24.13	0.2500	24.09	0.3490	0.1733	168	0.09	002.06	0
002	Protein, Crude (%)	0693	24.13	0.2500	24.09	0.3490	0.1733	168	0.09	002.06	0
002	Protein, Crude (%)	0968	24.13	0.0400	24.09	0.3490	0.1733	168	0.10	002.01	0
002	Protein, Crude (%)	0683	24.13	0.0800	24.09	0.3490	0.1733	168	0.10	002.05	0
002	Protein, Crude (%)	0038	24.13	0.0600	24.09	0.3490	0.1733	168	0.10	002.06	0
002	Protein, Crude (%)	0626	24.13	0.0820	24.09	0.3490	0.1733	168	0.11	002.06	0
002	Protein, Crude (%)	0074	24.14	0.5300	24.09	0.3490	0.1733	168	0.11	002.06	0
002	Protein, Crude (%)	0948	24.14	0.1200	24.09	0.3490	0.1733	168	0.13	002.05	0
002	Protein, Crude (%)	2009	24.15	0.0346	24.09	0.3490	0.1733	168	0.16	002.05	0
002	Protein, Crude (%)	0941	24.15	0.3400	24.09	0.3490	0.1733	168	0.16	002.05	0
002	Protein, Crude (%)	0366	24.15	0.1000	24.09	0.3490	0.1733	168	0.16	002.06	0
002	Protein, Crude (%)	0956	24.15	0.5000	24.09	0.3490	0.1733	168	0.16	002.06	0
002	Protein, Crude (%)	0089	24.16	0.0200	24.09	0.3490	0.1733	168	0.19	002.06	0
002	Protein, Crude (%)	0682	24.16	0.0000	24.09	0.3490	0.1733	168	0.19	002.06	0
002	Protein, Crude (%)	2184	24.17	0.1300	24.09	0.3490	0.1733	168	0.20	002.06	0
002	Protein, Crude (%)	0589	24.18	0.0500	24.09	0.3490	0.1733	168	0.23	002.06	0
002	Protein, Crude (%)	0571	24.18	0.0100	24.09	0.3490	0.1733	168	0.23	002.06	0
002	Protein, Crude (%)	0142	24.19	0.1100	24.09	0.3490	0.1733	168	0.26	002.06	0
002	Protein, Crude (%)	0914	24.19	0.0900	24.09	0.3490	0.1733	168	0.26	002.01	0
002	Protein, Crude (%)	0357	24.19	0.0000	24.09	0.3490	0.1733	168	0.27	002.06	0
002	Protein, Crude (%)	0164	24.19	0.1200	24.09	0.3490	0.1733	168	0.27	002.06	0
002	Protein, Crude (%)	0968	24.20	0.0300	24.09	0.3490	0.1733	168	0.29	002.06	0
002	Protein, Crude (%)	0045	24.20	0.0000	24.09	0.3490	0.1733	168	0.30	002.06	0
002	Protein, Crude (%)	0014	24.20	0.2000	24.09	0.3490	0.1733	168	0.30	002.06	0
002	Protein, Crude (%)	0098	24.20	0.2000	24.09	0.3490	0.1733	168	0.30	002.06	0
002	Protein, Crude (%)	2109	24.21	0.0300	24.09	0.3490	0.1733	168	0.32	002.06	0
002	Protein, Crude (%)	0674	24.21	0.0000	24.09	0.3490	0.1733	168	0.33	002.01	0
002	Protein, Crude (%)	0098	24.21	0.0800	24.09	0.3490	0.1733	168	0.33	002.08	0
002	Protein, Crude (%)	0687	24.22	0.0700	24.09	0.3490	0.1733	168	0.34	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0354	24.22	0.0200	24.09	0.3490	0.1733	168	0.36	002.06	0
002	Protein, Crude (%)	0553	24.23	0.1100	24.09	0.3490	0.1733	168	0.37	002.06	0
002	Protein, Crude (%)	2246	24.23	0.0200	24.09	0.3490	0.1733	168	0.39	002.06	0
002	Protein, Crude (%)	0723	24.25	0.0300	24.09	0.3490	0.1733	168	0.43	002.06	0
002	Protein, Crude (%)	0035	24.25	0.0200	24.09	0.3490	0.1733	168	0.44	002.06	0
002	Protein, Crude (%)	0407	24.25	0.0380	24.09	0.3490	0.1733	168	0.45	002.06	0
002	Protein, Crude (%)	0541	24.27	0.0100	24.09	0.3490	0.1733	168	0.49	002.06	0
002	Protein, Crude (%)	0939	24.27	0.0300	24.09	0.3490	0.1733	168	0.49	002.06	0
002	Protein, Crude (%)	0511	24.28	0.2300	24.09	0.3490	0.1733	168	0.52	002.06	0
002	Protein, Crude (%)	0508	24.28	0.1010	24.09	0.3490	0.1733	168	0.53	002.06	0
002	Protein, Crude (%)	2375	24.29	0.0900	24.09	0.3490	0.1733	168	0.54	002.06	0
002	Protein, Crude (%)	0610	24.30	0.2000	24.09	0.3490	0.1733	168	0.59	002.01	0
002	Protein, Crude (%)	2389	24.30	0.2000	24.09	0.3490	0.1733	168	0.59	002.06	0
002	Protein, Crude (%)	0123	24.30	0.0400	24.09	0.3490	0.1733	168	0.59	002.06	0
002	Protein, Crude (%)	2303	24.30	0.0400	24.09	0.3490	0.1733	168	0.59	002.06	0
002	Protein, Crude (%)	0358	24.34	0.5200	24.09	0.3490	0.1733	168	0.70	002.06	0
002	Protein, Crude (%)	0017	24.35	0.3000	24.09	0.3490	0.1733	168	0.73	002.06	0
002	Protein, Crude (%)	0171	24.39	0.0600	24.09	0.3490	0.1733	168	0.85	002.06	0
002	Protein, Crude (%)	0650	24.40	0.4000	24.09	0.3490	0.1733	168	0.87	002.06	0
002	Protein, Crude (%)	2302	24.41	0.2500	24.09	0.3490	0.1733	168	0.89	002.06	0
002	Protein, Crude (%)	0876	24.45	0.1040	24.09	0.3490	0.1733	168	1.01	002.06	0
002	Protein, Crude (%)	0656	24.45	0.0000	24.09	0.3490	0.1733	168	1.02	002.06	0
002	Protein, Crude (%)	0242	24.46	0.7000	24.09	0.3490	0.1733	168	1.05	002.06	0
002	Protein, Crude (%)	0941	24.47	0.1200	24.09	0.3490	0.1733	168	1.07	002.06	0
002	Protein, Crude (%)	0353	24.48	0.1900	24.09	0.3490	0.1733	168	1.09	002.06	0
002	Protein, Crude (%)	0190	24.48	0.0000	24.09	0.3490	0.1733	168	1.10	002.06	0
002	Protein, Crude (%)	0034	24.49	0.1400	24.09	0.3490	0.1733	168	1.13	002.06	0
002	Protein, Crude (%)	0512	24.49	0.0400	24.09	0.3490	0.1733	168	1.13	002.06	0
002	Protein, Crude (%)	0309	24.50	0.0495	24.09	0.3490	0.1733	168	1.17	002.06	0
002	Protein, Crude (%)	0425	24.53	0.0600	24.09	0.3490	0.1733	168	1.25	002.06	0
002	Protein, Crude (%)	0574	24.53	0.2400	24.09	0.3490	0.1733	168	1.25	002.06	0
002	Protein, Crude (%)	2359	24.55	0.3000	24.09	0.3490	0.1733	168	1.30	002.06	0
002	Protein, Crude (%)	2076	24.55	0.1290	24.09	0.3490	0.1733	168	1.31	002.06	0
002	Protein, Crude (%)	0918	24.56	0.4700	24.09	0.3490	0.1733	168	1.32	002.06	0
002	Protein, Crude (%)	0018	24.57	0.2600	24.09	0.3490	0.1733	168	1.36	002.06	0
002	Protein, Crude (%)	0229	24.57	0.8400	24.09	0.3490	0.1733	168	1.36	002.06	0
002	Protein, Crude (%)	0843	24.57	0.2800	24.09	0.3490	0.1733	168	1.36	002.06	0
002	Protein, Crude (%)	0001	24.68	0.3700	24.09	0.3490	0.1733	168	1.66	002.06	0
002	Protein, Crude (%)	0083	24.68	0.1180	24.09	0.3490	0.1733	168	1.67	002.06	0
002	Protein, Crude (%)	0226	24.69	0.1100	24.09	0.3490	0.1733	168	1.69	002.06	0
002	Protein, Crude (%)	2290	24.69	0.0100	24.09	0.3490	0.1733	168	1.69	002.11	0
002	Protein, Crude (%)	2089	24.70	1.280	24.09	0.3490	0.1733	168	1.73	002.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
002	Protein, Crude (%)	0656	24.71	0.0000	24.09	0.3490	0.1733	168	1.76	002.01	0
002	Protein, Crude (%)	2129	24.82	0.0570	24.09	0.3490	0.1733	168	2.07	002.06	0
002	Protein, Crude (%)	2268	24.84	0.5800	24.09	0.3490	0.1733	168	2.14	002.06	0
002	Protein, Crude (%)	0265	24.85	0.1000	24.09	0.3490	0.1733	168	2.16	002.06	0
002	Protein, Crude (%)	0912	24.90	0.0000	24.09	0.3490	0.1733	168	2.31	002.06	0
002	Protein, Crude (%)	0186	24.98	0.0400	24.09	0.3490	0.1733	168	2.54	002.06	0
002	Protein, Crude (%)	0618	25.11	0.1730	24.09	0.3490	0.1733	168	2.90	002.06	0
002	Protein, Crude (%)	2081	25.13	1.070	24.09	0.3490	0.1733	168	2.95	002.06	0
002	Protein, Crude (%)	2213	25.16	0.6600	24.09	0.3490	0.1733	168	3.05	002.06	0
002	Protein, Crude (%)	2326	26.16	0.0600	24.09	0.3490	0.1733	168	5.92	002.06	0
002	Protein, Crude (%)	0638	14.45	18.90	24.09	0.3490	0.1733	168	-27.64	002.04	2
003	Fat, Crude (%)	0357	2.555	0.0100	3.700	0.2690	0.1099	137	-4.26	003.14	0
003	Fat, Crude (%)	2089	2.810	0.8200	3.700	0.2690	0.1099	137	-3.31	003.13	0
003	Fat, Crude (%)	2359	3.000	0.0000	3.700	0.2690	0.1099	137	-2.60	003.14	0
003	Fat, Crude (%)	0939	3.015	0.2500	3.700	0.2690	0.1099	137	-2.55	003.14	0
003	Fat, Crude (%)	2246	3.035	0.0500	3.700	0.2690	0.1099	137	-2.47	003.13	0
003	Fat, Crude (%)	2302	3.145	0.0900	3.700	0.2690	0.1099	137	-2.06	003.14	0
003	Fat, Crude (%)	2073	3.155	0.0500	3.700	0.2690	0.1099	137	-2.03	003.10	0
003	Fat, Crude (%)	2022	3.200	0.1800	3.700	0.2690	0.1099	137	-1.86	003.10	0
003	Fat, Crude (%)	0265	3.220	0.1000	3.700	0.2690	0.1099	137	-1.78	003.14	0
003	Fat, Crude (%)	0242	3.230	0.2400	3.700	0.2690	0.1099	137	-1.75	003.10	0
003	Fat, Crude (%)	0042	3.315	0.1700	3.700	0.2690	0.1099	137	-1.43	003.10	0
003	Fat, Crude (%)	2302	3.330	0.0000	3.700	0.2690	0.1099	137	-1.38	003.11	0
003	Fat, Crude (%)	0142	3.335	0.2100	3.700	0.2690	0.1099	137	-1.36	003.00	0
003	Fat, Crude (%)	0144	3.350	0.0800	3.700	0.2690	0.1099	137	-1.30	003.14	0
003	Fat, Crude (%)	0345	3.355	0.1100	3.700	0.2690	0.1099	137	-1.28	003.00	0
003	Fat, Crude (%)	0682	3.360	0.0000	3.700	0.2690	0.1099	137	-1.26	003.14	0
003	Fat, Crude (%)	0003	3.380	0.3000	3.700	0.2690	0.1099	137	-1.19	003.14	0
003	Fat, Crude (%)	0536	3.380	0.1800	3.700	0.2690	0.1099	137	-1.19	003.99	0
003	Fat, Crude (%)	0876	3.387	0.0304	3.700	0.2690	0.1099	137	-1.16	003.14	0
003	Fat, Crude (%)	2144	3.405	0.3100	3.700	0.2690	0.1099	137	-1.10	003.14	0
003	Fat, Crude (%)	0309	3.410	0.0613	3.700	0.2690	0.1099	137	-1.08	003.00	0
003	Fat, Crude (%)	0693	3.425	0.1500	3.700	0.2690	0.1099	137	-1.02	003.10	0
003	Fat, Crude (%)	0504	3.450	0.8600	3.700	0.2690	0.1099	137	-0.93	003.01	0
003	Fat, Crude (%)	0674	3.450	0.0000	3.700	0.2690	0.1099	137	-0.93	003.06	0
003	Fat, Crude (%)	0505	3.455	0.0100	3.700	0.2690	0.1099	137	-0.91	003.14	0
003	Fat, Crude (%)	0676	3.485	0.1300	3.700	0.2690	0.1099	137	-0.80	003.10	0
003	Fat, Crude (%)	0581	3.485	0.0500	3.700	0.2690	0.1099	137	-0.80	003.14	0
003	Fat, Crude (%)	0036	3.501	0.0045	3.700	0.2690	0.1099	137	-0.74	003.10	0
003	Fat, Crude (%)	0619	3.505	0.0300	3.700	0.2690	0.1099	137	-0.72	003.10	0
003	Fat, Crude (%)	0066	3.520	0.0400	3.700	0.2690	0.1099	137	-0.67	003.14	0
003	Fat, Crude (%)	0541	3.525	0.0500	3.700	0.2690	0.1099	137	-0.65	003.10	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
003	Fat, Crude (%)	2340	3.529	0.0050	3.700	0.2690	0.1099	137	-0.64	003.06	0
003	Fat, Crude (%)	0169	3.540	0.0200	3.700	0.2690	0.1099	137	-0.59	003.06	0
003	Fat, Crude (%)	0045	3.540	0.0400	3.700	0.2690	0.1099	137	-0.59	003.10	0
003	Fat, Crude (%)	0004	3.545	0.1500	3.700	0.2690	0.1099	137	-0.58	003.09	0
003	Fat, Crude (%)	0278	3.550	0.1000	3.700	0.2690	0.1099	137	-0.56	003.14	0
003	Fat, Crude (%)	0956	3.550	0.1000	3.700	0.2690	0.1099	137	-0.56	003.14	0
003	Fat, Crude (%)	0263	3.552	0.0030	3.700	0.2690	0.1099	137	-0.55	003.10	0
003	Fat, Crude (%)	0912	3.565	0.1100	3.700	0.2690	0.1099	137	-0.50	003.14	0
003	Fat, Crude (%)	0358	3.570	0.0400	3.700	0.2690	0.1099	137	-0.48	003.14	0
003	Fat, Crude (%)	0512	3.580	0.0560	3.700	0.2690	0.1099	137	-0.45	003.09	0
003	Fat, Crude (%)	0034	3.580	0.1000	3.700	0.2690	0.1099	137	-0.45	003.14	0
003	Fat, Crude (%)	0083	3.585	0.0300	3.700	0.2690	0.1099	137	-0.43	003.14	0
003	Fat, Crude (%)	0019	3.595	0.0900	3.700	0.2690	0.1099	137	-0.39	003.14	0
003	Fat, Crude (%)	0848	3.595	0.0300	3.700	0.2690	0.1099	137	-0.39	003.14	0
003	Fat, Crude (%)	2375	3.595	0.0100	3.700	0.2690	0.1099	137	-0.39	003.14	0
003	Fat, Crude (%)	0366	3.600	0.4000	3.700	0.2690	0.1099	137	-0.37	003.10	0
003	Fat, Crude (%)	0297	3.600	0.0400	3.700	0.2690	0.1099	137	-0.37	003.11	0
003	Fat, Crude (%)	0089	3.610	0.0200	3.700	0.2690	0.1099	137	-0.33	003.14	0
003	Fat, Crude (%)	2326	3.615	0.2300	3.700	0.2690	0.1099	137	-0.32	003.12	0
003	Fat, Crude (%)	0108	3.615	0.0300	3.700	0.2690	0.1099	137	-0.32	003.14	0
003	Fat, Crude (%)	2155	3.619	0.0630	3.700	0.2690	0.1099	137	-0.30	003.10	0
003	Fat, Crude (%)	0004	3.620	0.0200	3.700	0.2690	0.1099	137	-0.30	003.14	0
003	Fat, Crude (%)	2196	3.630	0.0000	3.700	0.2690	0.1099	137	-0.26	003.10	0
003	Fat, Crude (%)	0202	3.630	0.1200	3.700	0.2690	0.1099	137	-0.26	003.14	0
003	Fat, Crude (%)	2192	3.635	0.0300	3.700	0.2690	0.1099	137	-0.24	003.00	0
003	Fat, Crude (%)	0062	3.640	0.0000	3.700	0.2690	0.1099	137	-0.22	003.10	0
003	Fat, Crude (%)	2188	3.640	0.0000	3.700	0.2690	0.1099	137	-0.22	003.10	0
003	Fat, Crude (%)	0098	3.645	0.1500	3.700	0.2690	0.1099	137	-0.20	003.10	0
003	Fat, Crude (%)	0171	3.650	0.0200	3.700	0.2690	0.1099	137	-0.19	003.14	0
003	Fat, Crude (%)	2397	3.655	0.0500	3.700	0.2690	0.1099	137	-0.17	003.10	0
003	Fat, Crude (%)	0651	3.664	0.6201	3.700	0.2690	0.1099	137	-0.13	003.06	0
003	Fat, Crude (%)	0968	3.665	0.0100	3.700	0.2690	0.1099	137	-0.13	003.14	0
003	Fat, Crude (%)	0425	3.670	0.0200	3.700	0.2690	0.1099	137	-0.11	003.06	0
003	Fat, Crude (%)	0885	3.670	0.0000	3.700	0.2690	0.1099	137	-0.11	003.11	0
003	Fat, Crude (%)	0148	3.675	0.0700	3.700	0.2690	0.1099	137	-0.09	003.14	0
003	Fat, Crude (%)	0123	3.685	0.0100	3.700	0.2690	0.1099	137	-0.06	003.10	0
003	Fat, Crude (%)	0100	3.690	0.1400	3.700	0.2690	0.1099	137	-0.04	003.10	0
003	Fat, Crude (%)	0885	3.694	0.1209	3.700	0.2690	0.1099	137	-0.02	003.00	0
003	Fat, Crude (%)	2190	3.700	0.1400	3.700	0.2690	0.1099	137	0.00	003.00	0
003	Fat, Crude (%)	0689	3.700	0.0000	3.700	0.2690	0.1099	137	0.00	003.06	0
003	Fat, Crude (%)	0689	3.700	0.0000	3.700	0.2690	0.1099	137	0.00	003.10	0
003	Fat, Crude (%)	0028	3.700	0.0000	3.700	0.2690	0.1099	137	0.00	003.13	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
003	Fat, Crude (%)	0229	3.705	0.2300	3.700	0.2690	0.1099	137	0.02	003.06	0
003	Fat, Crude (%)	0948	3.705	0.0300	3.700	0.2690	0.1099	137	0.02	003.14	0
003	Fat, Crude (%)	0033	3.725	0.0300	3.700	0.2690	0.1099	137	0.09	003.14	0
003	Fat, Crude (%)	0618	3.727	0.2340	3.700	0.2690	0.1099	137	0.10	003.10	0
003	Fat, Crude (%)	0629	3.730	0.0000	3.700	0.2690	0.1099	137	0.11	003.10	0
003	Fat, Crude (%)	0098	3.730	0.0200	3.700	0.2690	0.1099	137	0.11	003.13	0
003	Fat, Crude (%)	0035	3.730	0.0200	3.700	0.2690	0.1099	137	0.11	003.14	0
003	Fat, Crude (%)	0407	3.738	0.0139	3.700	0.2690	0.1099	137	0.14	003.14	0
003	Fat, Crude (%)	2006	3.750	0.1200	3.700	0.2690	0.1099	137	0.19	003.13	0
003	Fat, Crude (%)	2109	3.760	0.0600	3.700	0.2690	0.1099	137	0.22	003.00	0
003	Fat, Crude (%)	0148	3.765	0.0100	3.700	0.2690	0.1099	137	0.24	003.10	0
003	Fat, Crude (%)	0598	3.770	0.0200	3.700	0.2690	0.1099	137	0.26	003.14	0
003	Fat, Crude (%)	2184	3.770	0.1000	3.700	0.2690	0.1099	137	0.26	003.10	0
003	Fat, Crude (%)	2199	3.770	0.0600	3.700	0.2690	0.1099	137	0.26	003.12	0
003	Fat, Crude (%)	0529	3.775	0.1500	3.700	0.2690	0.1099	137	0.28	003.14	0
003	Fat, Crude (%)	0870	3.780	0.0171	3.700	0.2690	0.1099	137	0.30	003.13	0
003	Fat, Crude (%)	0675	3.785	0.0100	3.700	0.2690	0.1099	137	0.32	003.14	0
003	Fat, Crude (%)	0027	3.789	0.0260	3.700	0.2690	0.1099	137	0.33	003.09	0
003	Fat, Crude (%)	2076	3.789	0.2834	3.700	0.2690	0.1099	137	0.33	003.14	0
003	Fat, Crude (%)	0009	3.800	0.1400	3.700	0.2690	0.1099	137	0.37	003.14	0
003	Fat, Crude (%)	0692	3.800	0.6000	3.700	0.2690	0.1099	137	0.37	003.14	0
003	Fat, Crude (%)	0074	3.805	0.3700	3.700	0.2690	0.1099	137	0.39	003.14	0
003	Fat, Crude (%)	0870	3.809	0.1447	3.700	0.2690	0.1099	137	0.41	003.10	0
003	Fat, Crude (%)	2129	3.822	0.0710	3.700	0.2690	0.1099	137	0.45	003.14	0
003	Fat, Crude (%)	2349	3.835	0.1100	3.700	0.2690	0.1099	137	0.50	003.09	0
003	Fat, Crude (%)	0723	3.855	0.0300	3.700	0.2690	0.1099	137	0.58	003.09	0
003	Fat, Crude (%)	0294	3.865	0.0100	3.700	0.2690	0.1099	137	0.61	003.06	0
003	Fat, Crude (%)	0948	3.865	0.0700	3.700	0.2690	0.1099	137	0.61	003.09	0
003	Fat, Crude (%)	0186	3.870	0.0400	3.700	0.2690	0.1099	137	0.63	003.14	0
003	Fat, Crude (%)	0596	3.870	0.2200	3.700	0.2690	0.1099	137	0.63	003.14	0
003	Fat, Crude (%)	0098	3.875	0.0300	3.700	0.2690	0.1099	137	0.65	003.09	0
003	Fat, Crude (%)	0619	3.890	0.0800	3.700	0.2690	0.1099	137	0.71	003.14	0
003	Fat, Crude (%)	0910	3.895	0.0900	3.700	0.2690	0.1099	137	0.73	003.06	0
003	Fat, Crude (%)	2012	3.900	0.3000	3.700	0.2690	0.1099	137	0.74	003.09	0
003	Fat, Crude (%)	2319	3.900	0.0000	3.700	0.2690	0.1099	137	0.74	003.10	0
003	Fat, Crude (%)	2200	3.900	0.0000	3.700	0.2690	0.1099	137	0.74	003.12	0
003	Fat, Crude (%)	0968	3.910	0.1200	3.700	0.2690	0.1099	137	0.78	003.06	0
003	Fat, Crude (%)	0297	3.935	0.0700	3.700	0.2690	0.1099	137	0.87	003.06	0
003	Fat, Crude (%)	0574	3.940	0.0800	3.700	0.2690	0.1099	137	0.89	003.14	0
003	Fat, Crude (%)	2359	3.950	0.1000	3.700	0.2690	0.1099	137	0.93	003.00	0
003	Fat, Crude (%)	0610	3.950	0.0600	3.700	0.2690	0.1099	137	0.93	003.10	0
003	Fat, Crude (%)	0511	3.955	0.1700	3.700	0.2690	0.1099	137	0.95	003.14	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
003	Fat, Crude (%)	0354	3.960	0.0000	3.700	0.2690	0.1099	137	0.97	003.09	0
003	Fat, Crude (%)	0001	3.987	0.2860	3.700	0.2690	0.1099	137	1.07	003.14	0
003	Fat, Crude (%)	0017	4.000	0.2000	3.700	0.2690	0.1099	137	1.12	003.00	0
003	Fat, Crude (%)	0656	4.010	0.0000	3.700	0.2690	0.1099	137	1.15	003.99	0
003	Fat, Crude (%)	0941	4.010	0.1400	3.700	0.2690	0.1099	137	1.15	003.99	0
003	Fat, Crude (%)	0520	4.020	0.3600	3.700	0.2690	0.1099	137	1.19	003.14	0
003	Fat, Crude (%)	0229	4.050	0.0400	3.700	0.2690	0.1099	137	1.30	003.14	0
003	Fat, Crude (%)	0226	4.055	0.1500	3.700	0.2690	0.1099	137	1.32	003.14	0
003	Fat, Crude (%)	0687	4.100	0.2000	3.700	0.2690	0.1099	137	1.49	003.06	0
003	Fat, Crude (%)	0510	4.120	0.3200	3.700	0.2690	0.1099	137	1.56	003.14	0
003	Fat, Crude (%)	0300	4.125	0.1300	3.700	0.2690	0.1099	137	1.58	003.14	0
003	Fat, Crude (%)	2290	4.150	0.1000	3.700	0.2690	0.1099	137	1.67	003.11	0
003	Fat, Crude (%)	0644	4.177	0.0220	3.700	0.2690	0.1099	137	1.77	003.06	0
003	Fat, Crude (%)	0164	4.190	0.0600	3.700	0.2690	0.1099	137	1.82	003.01	0
003	Fat, Crude (%)	0682	4.210	0.0000	3.700	0.2690	0.1099	137	1.90	003.06	0
003	Fat, Crude (%)	2146	4.215	0.1500	3.700	0.2690	0.1099	137	1.92	003.14	0
003	Fat, Crude (%)	0726	4.245	0.0100	3.700	0.2690	0.1099	137	2.03	003.14	0
003	Fat, Crude (%)	0049	4.300	0.2200	3.700	0.2690	0.1099	137	2.23	003.14	0
003	Fat, Crude (%)	2325	4.315	0.0300	3.700	0.2690	0.1099	137	2.29	003.99	0
003	Fat, Crude (%)	0353	4.740	0.4800	3.700	0.2690	0.1099	137	3.87	003.09	0
003	Fat, Crude (%)	0190	4.790	0.0000	3.700	0.2690	0.1099	137	4.05	003.14	0
003	Fat, Crude (%)	0171	5.285	0.0700	3.700	0.2690	0.1099	137	5.89	003.12	0
003	Fat, Crude (%)	0638	6.700	6.600	3.700	0.2690	0.1099	137	11.15	003.09	1
003	Fat, Crude (%)	2337	10.41	0.0400	3.700	0.2690	0.1099	137	24.95	003.00	2
004	Fiber, Crude (%)	2199	4.600	0.2000	5.846	0.4626	0.1584	113	-2.69	004.06	0
004	Fiber, Crude (%)	0144	4.720	0.0400	5.846	0.4626	0.1584	113	-2.43	004.07	0
004	Fiber, Crude (%)	0353	4.750	0.1000	5.846	0.4626	0.1584	113	-2.37	004.03	0
004	Fiber, Crude (%)	2012	4.805	0.1300	5.846	0.4626	0.1584	113	-2.25	004.07	0
004	Fiber, Crude (%)	0510	4.900	0.0000	5.846	0.4626	0.1584	113	-2.05	004.07	0
004	Fiber, Crude (%)	0265	4.905	0.1100	5.846	0.4626	0.1584	113	-2.03	004.07	0
004	Fiber, Crude (%)	0366	5.100	0.2000	5.846	0.4626	0.1584	113	-1.61	004.07	0
004	Fiber, Crude (%)	0956	5.200	0.4000	5.846	0.4626	0.1584	113	-1.40	004.07	0
004	Fiber, Crude (%)	0629	5.200	0.0000	5.846	0.4626	0.1584	113	-1.40	004.99	0
004	Fiber, Crude (%)	2190	5.210	0.3000	5.846	0.4626	0.1584	113	-1.37	004.07	0
004	Fiber, Crude (%)	0407	5.220	0.1501	5.846	0.4626	0.1584	113	-1.35	004.07	0
004	Fiber, Crude (%)	2319	5.250	0.1000	5.846	0.4626	0.1584	113	-1.29	004.06	0
004	Fiber, Crude (%)	0644	5.264	0.0260	5.846	0.4626	0.1584	113	-1.26	004.07	0
004	Fiber, Crude (%)	0294	5.275	0.4500	5.846	0.4626	0.1584	113	-1.23	004.07	0
004	Fiber, Crude (%)	0028	5.300	0.2000	5.846	0.4626	0.1584	113	-1.18	004.07	0
004	Fiber, Crude (%)	0098	5.425	0.4900	5.846	0.4626	0.1584	113	-0.91	004.07	0
004	Fiber, Crude (%)	0708	5.430	0.1200	5.846	0.4626	0.1584	113	-0.90	004.07	0
004	Fiber, Crude (%)	2397	5.450	0.3200	5.846	0.4626	0.1584	113	-0.86	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
004	Fiber, Crude (%)	0169	5.465	0.0500	5.846	0.4626	0.1584	113	-0.82	004.00	0
004	Fiber, Crude (%)	2200	5.500	0.2000	5.846	0.4626	0.1584	113	-0.75	004.06	0
004	Fiber, Crude (%)	2022	5.510	0.2400	5.846	0.4626	0.1584	113	-0.73	004.07	0
004	Fiber, Crude (%)	0309	5.513	0.1058	5.846	0.4626	0.1584	113	-0.72	004.00	0
004	Fiber, Crude (%)	0683	5.520	0.0400	5.846	0.4626	0.1584	113	-0.70	004.07	0
004	Fiber, Crude (%)	2009	5.540	0.0120	5.846	0.4626	0.1584	113	-0.66	004.07	0
004	Fiber, Crude (%)	0589	5.570	0.4400	5.846	0.4626	0.1584	113	-0.60	004.07	0
004	Fiber, Crude (%)	0038	5.580	0.0800	5.846	0.4626	0.1584	113	-0.58	004.06	0
004	Fiber, Crude (%)	2302	5.585	0.0100	5.846	0.4626	0.1584	113	-0.56	004.07	0
004	Fiber, Crude (%)	0656	5.590	0.0000	5.846	0.4626	0.1584	113	-0.55	004.07	0
004	Fiber, Crude (%)	0619	5.600	0.0800	5.846	0.4626	0.1584	113	-0.53	004.07	0
004	Fiber, Crude (%)	0682	5.600	0.0000	5.846	0.4626	0.1584	113	-0.53	004.07	0
004	Fiber, Crude (%)	0100	5.615	0.2300	5.846	0.4626	0.1584	113	-0.50	004.07	0
004	Fiber, Crude (%)	0035	5.625	0.0100	5.846	0.4626	0.1584	113	-0.48	004.07	0
004	Fiber, Crude (%)	2146	5.630	0.1400	5.846	0.4626	0.1584	113	-0.47	004.00	0
004	Fiber, Crude (%)	0610	5.640	0.0800	5.846	0.4626	0.1584	113	-0.45	004.06	0
004	Fiber, Crude (%)	0009	5.640	0.1400	5.846	0.4626	0.1584	113	-0.45	004.07	0
004	Fiber, Crude (%)	0968	5.640	0.1000	5.846	0.4626	0.1584	113	-0.45	004.07	0
004	Fiber, Crude (%)	0726	5.645	0.0100	5.846	0.4626	0.1584	113	-0.43	004.07	0
004	Fiber, Crude (%)	0910	5.650	0.3000	5.846	0.4626	0.1584	113	-0.42	004.07	0
004	Fiber, Crude (%)	0186	5.655	0.0900	5.846	0.4626	0.1584	113	-0.41	004.00	0
004	Fiber, Crude (%)	0083	5.670	0.2600	5.846	0.4626	0.1584	113	-0.38	004.07	0
004	Fiber, Crude (%)	0300	5.670	0.0600	5.846	0.4626	0.1584	113	-0.38	004.07	0
004	Fiber, Crude (%)	2337	5.690	0.0400	5.846	0.4626	0.1584	113	-0.34	004.07	0
004	Fiber, Crude (%)	0675	5.695	0.0700	5.846	0.4626	0.1584	113	-0.33	004.07	0
004	Fiber, Crude (%)	0164	5.700	0.4000	5.846	0.4626	0.1584	113	-0.32	004.00	0
004	Fiber, Crude (%)	0692	5.700	0.0000	5.846	0.4626	0.1584	113	-0.32	004.07	0
004	Fiber, Crude (%)	0123	5.715	0.3900	5.846	0.4626	0.1584	113	-0.28	004.07	0
004	Fiber, Crude (%)	0242	5.720	0.2400	5.846	0.4626	0.1584	113	-0.27	004.07	0
004	Fiber, Crude (%)	2340	5.735	0.0040	5.846	0.4626	0.1584	113	-0.24	004.07	0
004	Fiber, Crude (%)	0505	5.745	0.0700	5.846	0.4626	0.1584	113	-0.22	004.07	0
004	Fiber, Crude (%)	0870	5.746	0.0143	5.846	0.4626	0.1584	113	-0.22	004.07	0
004	Fiber, Crude (%)	0425	5.750	0.1000	5.846	0.4626	0.1584	113	-0.21	004.00	0
004	Fiber, Crude (%)	2349	5.750	0.3000	5.846	0.4626	0.1584	113	-0.21	004.06	0
004	Fiber, Crude (%)	2188	5.760	0.0000	5.846	0.4626	0.1584	113	-0.19	004.06	0
004	Fiber, Crude (%)	0066	5.780	0.1400	5.846	0.4626	0.1584	113	-0.14	004.07	0
004	Fiber, Crude (%)	0553	5.785	0.1300	5.846	0.4626	0.1584	113	-0.13	004.07	0
004	Fiber, Crude (%)	0098	5.795	0.0100	5.846	0.4626	0.1584	113	-0.11	004.06	0
004	Fiber, Crude (%)	0689	5.800	0.4000	5.846	0.4626	0.1584	113	-0.10	004.06	0
004	Fiber, Crude (%)	0939	5.815	0.1900	5.846	0.4626	0.1584	113	-0.07	004.07	0
004	Fiber, Crude (%)	2144	5.820	0.1000	5.846	0.4626	0.1584	113	-0.06	004.07	0
004	Fiber, Crude (%)	0643	5.825	0.0300	5.846	0.4626	0.1584	113	-0.05	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
004	Fiber, Crude (%)	0723	5.830	0.0400	5.846	0.4626	0.1584	113	-0.03	004.07	0
004	Fiber, Crude (%)	0017	5.850	0.1000	5.846	0.4626	0.1584	113	0.01	004.00	0
004	Fiber, Crude (%)	0511	5.850	0.1000	5.846	0.4626	0.1584	113	0.01	004.07	0
004	Fiber, Crude (%)	0674	5.860	0.0000	5.846	0.4626	0.1584	113	0.03	004.06	0
004	Fiber, Crude (%)	0512	5.880	0.4300	5.846	0.4626	0.1584	113	0.07	004.07	0
004	Fiber, Crude (%)	2340	5.885	0.0350	5.846	0.4626	0.1584	113	0.08	004.06	0
004	Fiber, Crude (%)	2146	5.890	0.1000	5.846	0.4626	0.1584	113	0.09	004.07	0
004	Fiber, Crude (%)	2006	5.905	0.1500	5.846	0.4626	0.1584	113	0.13	004.06	0
004	Fiber, Crude (%)	0019	5.915	0.1300	5.846	0.4626	0.1584	113	0.15	004.07	0
004	Fiber, Crude (%)	2196	5.920	0.0000	5.846	0.4626	0.1584	113	0.16	004.07	0
004	Fiber, Crude (%)	0885	5.925	0.1469	5.846	0.4626	0.1584	113	0.17	004.07	0
004	Fiber, Crude (%)	0027	5.942	0.0150	5.846	0.4626	0.1584	113	0.21	004.06	0
004	Fiber, Crude (%)	2089	5.950	0.5400	5.846	0.4626	0.1584	113	0.22	004.03	0
004	Fiber, Crude (%)	2129	5.958	0.1160	5.846	0.4626	0.1584	113	0.24	004.99	0
004	Fiber, Crude (%)	0354	5.965	0.2900	5.846	0.4626	0.1584	113	0.26	004.07	0
004	Fiber, Crude (%)	0345	5.970	0.1600	5.846	0.4626	0.1584	113	0.27	004.00	0
004	Fiber, Crude (%)	0968	5.990	0.2200	5.846	0.4626	0.1584	113	0.31	004.06	0
004	Fiber, Crude (%)	2359	6.000	0.0000	5.846	0.4626	0.1584	113	0.33	004.00	0
004	Fiber, Crude (%)	2259	6.014	0.0930	5.846	0.4626	0.1584	113	0.36	004.07	0
004	Fiber, Crude (%)	0263	6.019	0.1510	5.846	0.4626	0.1584	113	0.37	004.07	0
004	Fiber, Crude (%)	0876	6.034	0.0458	5.846	0.4626	0.1584	113	0.41	004.07	0
004	Fiber, Crude (%)	2290	6.050	0.1000	5.846	0.4626	0.1584	113	0.44	004.11	0
004	Fiber, Crude (%)	2022	6.055	0.0900	5.846	0.4626	0.1584	113	0.45	004.06	0
004	Fiber, Crude (%)	0520	6.075	0.0100	5.846	0.4626	0.1584	113	0.49	004.07	0
004	Fiber, Crude (%)	0034	6.090	0.1600	5.846	0.4626	0.1584	113	0.53	004.07	0
004	Fiber, Crude (%)	0171	6.100	0.0200	5.846	0.4626	0.1584	113	0.55	004.00	0
004	Fiber, Crude (%)	0948	6.120	0.1800	5.846	0.4626	0.1584	113	0.59	004.06	0
004	Fiber, Crude (%)	2076	6.136	0.6490	5.846	0.4626	0.1584	113	0.63	004.00	0
004	Fiber, Crude (%)	0089	6.170	0.0000	5.846	0.4626	0.1584	113	0.70	004.07	0
004	Fiber, Crude (%)	0885	6.210	0.0400	5.846	0.4626	0.1584	113	0.79	004.11	0
004	Fiber, Crude (%)	0504	6.255	0.1300	5.846	0.4626	0.1584	113	0.88	004.07	0
004	Fiber, Crude (%)	0693	6.275	0.2700	5.846	0.4626	0.1584	113	0.93	004.07	0
004	Fiber, Crude (%)	2188	6.280	0.0000	5.846	0.4626	0.1584	113	0.94	004.07	0
004	Fiber, Crude (%)	0190	6.300	0.0000	5.846	0.4626	0.1584	113	0.98	004.07	0
004	Fiber, Crude (%)	0074	6.320	0.2000	5.846	0.4626	0.1584	113	1.02	004.07	0
004	Fiber, Crude (%)	0948	6.345	0.1500	5.846	0.4626	0.1584	113	1.08	004.07	0
004	Fiber, Crude (%)	0229	6.405	0.5500	5.846	0.4626	0.1584	113	1.21	004.07	0
004	Fiber, Crude (%)	0581	6.505	0.1100	5.846	0.4626	0.1584	113	1.42	004.07	0
004	Fiber, Crude (%)	0529	6.560	0.0600	5.846	0.4626	0.1584	113	1.54	004.07	0
004	Fiber, Crude (%)	0676	6.575	0.3300	5.846	0.4626	0.1584	113	1.58	004.06	0
004	Fiber, Crude (%)	0004	6.630	0.1200	5.846	0.4626	0.1584	113	1.69	004.07	0
004	Fiber, Crude (%)	0598	6.705	0.0300	5.846	0.4626	0.1584	113	1.86	004.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
004	Fiber, Crude (%)	2302	6.750	0.0000	5.846	0.4626	0.1584	113	1.95	004.11	0
004	Fiber, Crude (%)	2246	6.765	0.3500	5.846	0.4626	0.1584	113	1.99	004.03	0
004	Fiber, Crude (%)	0042	6.805	0.8100	5.846	0.4626	0.1584	113	2.07	004.07	0
004	Fiber, Crude (%)	2192	6.810	0.1800	5.846	0.4626	0.1584	113	2.08	004.03	0
004	Fiber, Crude (%)	0848	6.855	0.0500	5.846	0.4626	0.1584	113	2.18	004.07	0
004	Fiber, Crude (%)	0045	6.890	0.1400	5.846	0.4626	0.1584	113	2.26	004.07	0
004	Fiber, Crude (%)	2326	6.945	0.7100	5.846	0.4626	0.1584	113	2.38	004.00	0
004	Fiber, Crude (%)	0914	7.250	0.3000	5.846	0.4626	0.1584	113	3.03	004.07	0
004	Fiber, Crude (%)	0226	7.860	0.1600	5.846	0.4626	0.1584	113	4.35	004.07	0
004	Fiber, Crude (%)	0278	7.950	0.3000	5.846	0.4626	0.1584	113	4.55	004.07	0
004	Fiber, Crude (%)	0536	8.330	0.2000	5.846	0.4626	0.1584	113	5.37	004.07	0
004	Fiber, Crude (%)	0638	8.650	2.700	5.846	0.4626	0.1584	113	6.06	004.06	1
004	Fiber, Crude (%)	0626	11.58	0.0390	5.846	0.4626	0.1584	113	12.39	004.07	2
005	Ash (%)	2081	5.965	0.1100	7.106	0.2703	0.0829	134	-4.22	005.99	0
005	Ash (%)	0511	6.450	0.3000	7.106	0.2703	0.0829	134	-2.43	005.03	0
005	Ash (%)	0953	6.475	0.0900	7.106	0.2703	0.0829	134	-2.33	005.00	0
005	Ash (%)	2184	6.610	0.0400	7.106	0.2703	0.0829	134	-1.83	005.00	0
005	Ash (%)	0108	6.615	0.1300	7.106	0.2703	0.0829	134	-1.81	005.00	0
005	Ash (%)	0169	6.630	0.0200	7.106	0.2703	0.0829	134	-1.76	005.00	0
005	Ash (%)	0019	6.650	0.1000	7.106	0.2703	0.0829	134	-1.69	005.99	0
005	Ash (%)	2199	6.700	0.2000	7.106	0.2703	0.0829	134	-1.50	005.00	0
005	Ash (%)	0651	6.711	0.1216	7.106	0.2703	0.0829	134	-1.46	005.00	0
005	Ash (%)	0539	6.740	0.0600	7.106	0.2703	0.0829	134	-1.35	005.00	0
005	Ash (%)	0049	6.745	0.0500	7.106	0.2703	0.0829	134	-1.33	005.00	0
005	Ash (%)	0510	6.750	0.1400	7.106	0.2703	0.0829	134	-1.32	005.00	0
005	Ash (%)	0529	6.750	0.0000	7.106	0.2703	0.0829	134	-1.32	005.00	0
005	Ash (%)	0650	6.755	0.0700	7.106	0.2703	0.0829	134	-1.30	005.00	0
005	Ash (%)	0042	6.770	0.1200	7.106	0.2703	0.0829	134	-1.24	005.00	0
005	Ash (%)	0004	6.790	0.0400	7.106	0.2703	0.0829	134	-1.17	005.00	0
005	Ash (%)	0644	6.791	0.0270	7.106	0.2703	0.0829	134	-1.17	005.99	0
005	Ash (%)	2359	6.800	0.0000	7.106	0.2703	0.0829	134	-1.13	005.00	0
005	Ash (%)	2022	6.815	0.0300	7.106	0.2703	0.0829	134	-1.07	005.00	0
005	Ash (%)	0003	6.820	0.0200	7.106	0.2703	0.0829	134	-1.06	005.05	0
005	Ash (%)	0186	6.821	0.0936	7.106	0.2703	0.0829	134	-1.05	005.00	0
005	Ash (%)	0066	6.825	0.0500	7.106	0.2703	0.0829	134	-1.04	005.05	0
005	Ash (%)	0425	6.825	0.0100	7.106	0.2703	0.0829	134	-1.04	005.00	0
005	Ash (%)	2340	6.826	0.0310	7.106	0.2703	0.0829	134	-1.04	005.00	0
005	Ash (%)	2290	6.850	0.3000	7.106	0.2703	0.0829	134	-0.95	005.11	0
005	Ash (%)	0164	6.855	0.0900	7.106	0.2703	0.0829	134	-0.93	005.00	0
005	Ash (%)	0505	6.855	0.0900	7.106	0.2703	0.0829	134	-0.93	005.00	0
005	Ash (%)	0918	6.855	0.0300	7.106	0.2703	0.0829	134	-0.93	005.00	0
005	Ash (%)	2349	6.870	0.1600	7.106	0.2703	0.0829	134	-0.87	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
005	Ash (%)	0598	6.875	0.0100	7.106	0.2703	0.0829	134	-0.85	005.00	0
005	Ash (%)	0357	6.900	0.2000	7.106	0.2703	0.0829	134	-0.76	005.00	0
005	Ash (%)	2302	6.905	0.1300	7.106	0.2703	0.0829	134	-0.74	005.00	0
005	Ash (%)	0144	6.915	0.3900	7.106	0.2703	0.0829	134	-0.71	005.00	0
005	Ash (%)	2268	6.915	0.2700	7.106	0.2703	0.0829	134	-0.71	005.00	0
005	Ash (%)	0027	6.927	0.1110	7.106	0.2703	0.0829	134	-0.66	005.00	0
005	Ash (%)	0407	6.934	0.0453	7.106	0.2703	0.0829	134	-0.63	005.00	0
005	Ash (%)	2337	6.935	0.0100	7.106	0.2703	0.0829	134	-0.63	005.00	0
005	Ash (%)	0038	6.940	0.1400	7.106	0.2703	0.0829	134	-0.61	005.05	0
005	Ash (%)	0848	6.945	0.0300	7.106	0.2703	0.0829	134	-0.59	005.00	0
005	Ash (%)	0278	6.950	0.2200	7.106	0.2703	0.0829	134	-0.58	005.00	0
005	Ash (%)	0148	6.950	0.0800	7.106	0.2703	0.0829	134	-0.58	005.00	0
005	Ash (%)	0366	6.950	0.1000	7.106	0.2703	0.0829	134	-0.58	005.00	0
005	Ash (%)	0956	6.950	0.5000	7.106	0.2703	0.0829	134	-0.58	005.00	0
005	Ash (%)	0693	6.960	0.1600	7.106	0.2703	0.0829	134	-0.54	005.00	0
005	Ash (%)	0358	6.970	0.0200	7.106	0.2703	0.0829	134	-0.50	005.00	0
005	Ash (%)	2246	6.980	0.0000	7.106	0.2703	0.0829	134	-0.46	005.00	0
005	Ash (%)	2397	6.980	0.0600	7.106	0.2703	0.0829	134	-0.46	005.00	0
005	Ash (%)	0229	6.990	0.1200	7.106	0.2703	0.0829	134	-0.43	005.00	0
005	Ash (%)	2188	6.990	0.0000	7.106	0.2703	0.0829	134	-0.43	005.00	0
005	Ash (%)	0589	7.000	0.0400	7.106	0.2703	0.0829	134	-0.39	005.00	0
005	Ash (%)	0541	7.015	0.1500	7.106	0.2703	0.0829	134	-0.34	005.00	0
005	Ash (%)	0242	7.020	0.0200	7.106	0.2703	0.0829	134	-0.32	005.00	0
005	Ash (%)	0265	7.050	0.1000	7.106	0.2703	0.0829	134	-0.21	005.00	0
005	Ash (%)	2150	7.055	0.0300	7.106	0.2703	0.0829	134	-0.19	005.00	0
005	Ash (%)	0723	7.065	0.0100	7.106	0.2703	0.0829	134	-0.15	005.00	0
005	Ash (%)	0123	7.070	0.0600	7.106	0.2703	0.0829	134	-0.13	005.00	0
005	Ash (%)	2196	7.070	0.0000	7.106	0.2703	0.0829	134	-0.13	005.00	0
005	Ash (%)	0574	7.075	0.0700	7.106	0.2703	0.0829	134	-0.11	005.00	0
005	Ash (%)	0098	7.100	0.0200	7.106	0.2703	0.0829	134	-0.02	005.00	0
005	Ash (%)	2006	7.100	0.1200	7.106	0.2703	0.0829	134	-0.02	005.00	0
005	Ash (%)	0692	7.100	0.2000	7.106	0.2703	0.0829	134	-0.02	005.05	0
005	Ash (%)	0876	7.101	0.0814	7.106	0.2703	0.0829	134	-0.02	005.00	0
005	Ash (%)	0142	7.105	0.0100	7.106	0.2703	0.0829	134	0.00	005.00	0
005	Ash (%)	0643	7.105	0.0300	7.106	0.2703	0.0829	134	0.00	005.05	0
005	Ash (%)	0354	7.109	0.0220	7.106	0.2703	0.0829	134	0.01	005.00	0
005	Ash (%)	0553	7.120	0.2000	7.106	0.2703	0.0829	134	0.05	005.00	0
005	Ash (%)	0001	7.121	0.0830	7.106	0.2703	0.0829	134	0.06	005.00	0
005	Ash (%)	0726	7.125	0.0500	7.106	0.2703	0.0829	134	0.07	005.05	0
005	Ash (%)	0870	7.130	0.1232	7.106	0.2703	0.0829	134	0.09	005.00	0
005	Ash (%)	0100	7.130	0.1600	7.106	0.2703	0.0829	134	0.09	005.00	0
005	Ash (%)	0520	7.145	0.0900	7.106	0.2703	0.0829	134	0.15	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
005	Ash (%)	2200	7.150	0.1000	7.106	0.2703	0.0829	134	0.16	005.00	0
005	Ash (%)	0618	7.155	0.0300	7.106	0.2703	0.0829	134	0.18	005.00	0
005	Ash (%)	0689	7.155	0.0900	7.106	0.2703	0.0829	134	0.18	005.05	0
005	Ash (%)	0619	7.163	0.0110	7.106	0.2703	0.0829	134	0.21	005.05	0
005	Ash (%)	0035	7.165	0.0100	7.106	0.2703	0.0829	134	0.22	005.00	0
005	Ash (%)	0083	7.170	0.0200	7.106	0.2703	0.0829	134	0.24	005.00	0
005	Ash (%)	0504	7.170	0.3800	7.106	0.2703	0.0829	134	0.24	005.00	0
005	Ash (%)	0968	7.170	0.0200	7.106	0.2703	0.0829	134	0.24	005.00	0
005	Ash (%)	0171	7.185	0.0100	7.106	0.2703	0.0829	134	0.29	005.00	0
005	Ash (%)	2076	7.197	0.0042	7.106	0.2703	0.0829	134	0.34	005.00	0
005	Ash (%)	0047	7.200	0.2000	7.106	0.2703	0.0829	134	0.35	005.00	0
005	Ash (%)	0017	7.200	0.0000	7.106	0.2703	0.0829	134	0.35	005.00	0
005	Ash (%)	2193	7.200	0.0800	7.106	0.2703	0.0829	134	0.35	005.00	0
005	Ash (%)	0294	7.200	0.0000	7.106	0.2703	0.0829	134	0.35	005.05	0
005	Ash (%)	2192	7.200	0.0200	7.106	0.2703	0.0829	134	0.35	005.05	0
005	Ash (%)	2259	7.201	0.0220	7.106	0.2703	0.0829	134	0.35	005.00	0
005	Ash (%)	0309	7.202	0.0119	7.106	0.2703	0.0829	134	0.36	005.00	0
005	Ash (%)	2326	7.205	0.0900	7.106	0.2703	0.0829	134	0.37	005.00	0
005	Ash (%)	0297	7.210	0.1600	7.106	0.2703	0.0829	134	0.39	005.00	0
005	Ash (%)	0682	7.210	0.0000	7.106	0.2703	0.0829	134	0.39	005.00	0
005	Ash (%)	0202	7.210	0.1200	7.106	0.2703	0.0829	134	0.39	005.99	0
005	Ash (%)	0656	7.220	0.3600	7.106	0.2703	0.0829	134	0.42	005.00	0
005	Ash (%)	0353	7.235	0.1300	7.106	0.2703	0.0829	134	0.48	005.00	0
005	Ash (%)	0263	7.247	0.0020	7.106	0.2703	0.0829	134	0.52	005.05	0
005	Ash (%)	0683	7.250	0.0400	7.106	0.2703	0.0829	134	0.53	005.05	0
005	Ash (%)	0045	7.255	0.0500	7.106	0.2703	0.0829	134	0.55	005.00	0
005	Ash (%)	2325	7.300	0.0000	7.106	0.2703	0.0829	134	0.72	005.00	0
005	Ash (%)	2246	7.310	0.0400	7.106	0.2703	0.0829	134	0.76	005.05	0
005	Ash (%)	2340	7.312	0.0030	7.106	0.2703	0.0829	134	0.76	005.05	0
005	Ash (%)	0512	7.321	0.1710	7.106	0.2703	0.0829	134	0.79	005.00	0
005	Ash (%)	2144	7.325	0.0300	7.106	0.2703	0.0829	134	0.81	005.05	0
005	Ash (%)	2012	7.330	0.1600	7.106	0.2703	0.0829	134	0.83	005.05	0
005	Ash (%)	0536	7.330	0.0400	7.106	0.2703	0.0829	134	0.83	005.99	0
005	Ash (%)	2146	7.335	0.0900	7.106	0.2703	0.0829	134	0.85	005.05	0
005	Ash (%)	0596	7.345	0.0100	7.106	0.2703	0.0829	134	0.89	005.05	0
005	Ash (%)	2190	7.345	0.0700	7.106	0.2703	0.0829	134	0.89	005.05	0
005	Ash (%)	0912	7.350	0.1000	7.106	0.2703	0.0829	134	0.90	005.03	0
005	Ash (%)	0939	7.350	0.0800	7.106	0.2703	0.0829	134	0.90	005.05	0
005	Ash (%)	0941	7.350	0.0400	7.106	0.2703	0.0829	134	0.90	005.05	0
005	Ash (%)	2319	7.350	0.1000	7.106	0.2703	0.0829	134	0.90	005.99	0
005	Ash (%)	0062	7.360	0.1200	7.106	0.2703	0.0829	134	0.94	005.00	0
005	Ash (%)	0914	7.365	0.0300	7.106	0.2703	0.0829	134	0.96	005.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
005	Ash (%)	0948	7.365	0.0500	7.106	0.2703	0.0829	134	0.96	005.00	0
005	Ash (%)	0675	7.370	0.0000	7.106	0.2703	0.0829	134	0.98	005.00	0
005	Ash (%)	0629	7.370	0.0800	7.106	0.2703	0.0829	134	0.98	005.99	0
005	Ash (%)	0610	7.385	0.0500	7.106	0.2703	0.0829	134	1.03	005.05	0
005	Ash (%)	2109	7.385	0.1900	7.106	0.2703	0.0829	134	1.03	005.05	0
005	Ash (%)	2089	7.400	0.1400	7.106	0.2703	0.0829	134	1.09	005.00	0
005	Ash (%)	0676	7.400	0.0400	7.106	0.2703	0.0829	134	1.09	005.99	0
005	Ash (%)	0644	7.410	0.0180	7.106	0.2703	0.0829	134	1.13	005.02	0
005	Ash (%)	0885	7.417	0.0077	7.106	0.2703	0.0829	134	1.15	005.05	0
005	Ash (%)	0345	7.425	0.0300	7.106	0.2703	0.0829	134	1.18	005.00	0
005	Ash (%)	0910	7.450	0.1000	7.106	0.2703	0.0829	134	1.27	005.99	0
005	Ash (%)	0226	7.495	0.0700	7.106	0.2703	0.0829	134	1.44	005.00	0
005	Ash (%)	0034	7.498	0.0260	7.106	0.2703	0.0829	134	1.45	005.05	0
005	Ash (%)	0190	7.555	0.0100	7.106	0.2703	0.0829	134	1.66	005.05	0
005	Ash (%)	2129	7.581	0.1920	7.106	0.2703	0.0829	134	1.76	005.05	0
005	Ash (%)	0652	7.600	0.0000	7.106	0.2703	0.0829	134	1.83	005.99	0
005	Ash (%)	2302	7.690	0.0000	7.106	0.2703	0.0829	134	2.16	005.11	0
005	Ash (%)	0674	7.770	0.0000	7.106	0.2703	0.0829	134	2.46	005.00	0
005	Ash (%)	0885	8.475	0.0100	7.106	0.2703	0.0829	134	5.07	005.11	0
005	Ash (%)	2389	9.555	0.0300	7.106	0.2703	0.0829	134	9.06	005.11	0
005	Ash (%)	0297	10.32	0.2300	7.106	0.2703	0.0829	134	11.87	005.11	0
005	Ash (%)	0089	5.130	3.020	7.106	0.2703	0.0829	134	-7.31	005.00	1
005	Ash (%)	0638	5.650	3.300	7.106	0.2703	0.0829	134	-5.38	005.05	1
005	Ash (%)	2009	3.551	0.0544	7.106	0.2703	0.0829	134	-13.15	005.05	2
006	Total Sugars (%)	0148	3.500	0.1200	5.247	1.090	0.1340	13	-1.60	006.99	0
006	Total Sugars (%)	0918	3.510	0.1600	5.247	1.090	0.1340	13	-1.59	006.99	0
006	Total Sugars (%)	2192	4.355	0.4700	5.247	1.090	0.1340	13	-0.82	006.00	0
006	Total Sugars (%)	0910	4.845	0.3500	5.247	1.090	0.1340	13	-0.37	006.00	0
006	Total Sugars (%)	0674	4.930	0.0000	5.247	1.090	0.1340	13	-0.29	006.99	0
006	Total Sugars (%)	0407	5.258	0.0200	5.247	1.090	0.1340	13	0.01	006.00	0
006	Total Sugars (%)	2319	5.400	0.0000	5.247	1.090	0.1340	13	0.14	006.00	0
006	Total Sugars (%)	0610	5.630	0.0200	5.247	1.090	0.1340	13	0.35	006.99	0
006	Total Sugars (%)	0226	5.670	0.0000	5.247	1.090	0.1340	13	0.39	006.99	0
006	Total Sugars (%)	0885	5.846	0.1024	5.247	1.090	0.1340	13	0.55	006.00	0
006	Total Sugars (%)	2290	5.950	0.1000	5.247	1.090	0.1340	13	0.65	006.99	0
006	Total Sugars (%)	0956	6.350	0.3000	5.247	1.090	0.1340	13	1.01	006.99	0
006	Total Sugars (%)	0912	6.750	0.1000	5.247	1.090	0.1340	13	1.38	006.99	0
008	Fiber, Acid Detergent (%)	2193	5.670	0.1400	8.175	0.5355	0.2530	61	-4.68	008.08	0
008	Fiber, Acid Detergent (%)	2129	6.809	0.1830	8.175	0.5355	0.2530	61	-2.55	008.08	0
008	Fiber, Acid Detergent (%)	0345	7.000	0.2000	8.175	0.5355	0.2530	61	-2.19	008.02	0
008	Fiber, Acid Detergent (%)	0723	7.290	0.2000	8.175	0.5355	0.2530	61	-1.65	008.08	0
008	Fiber, Acid Detergent (%)	0956	7.350	0.7000	8.175	0.5355	0.2530	61	-1.54	008.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
008	Fiber, Acid Detergent (%)	0676	7.370	0.1400	8.175	0.5355	0.2530	61	-1.50	008.02	0
008	Fiber, Acid Detergent (%)	2302	7.510	0.0000	8.175	0.5355	0.2530	61	-1.24	008.08	0
008	Fiber, Acid Detergent (%)	0968	7.530	0.0200	8.175	0.5355	0.2530	61	-1.21	008.08	0
008	Fiber, Acid Detergent (%)	0675	7.580	0.1400	8.175	0.5355	0.2530	61	-1.11	008.08	0
008	Fiber, Acid Detergent (%)	0885	7.682	0.3670	8.175	0.5355	0.2530	61	-0.92	008.08	0
008	Fiber, Acid Detergent (%)	0098	7.735	0.1300	8.175	0.5355	0.2530	61	-0.82	008.02	0
008	Fiber, Acid Detergent (%)	0358	7.745	0.0100	8.175	0.5355	0.2530	61	-0.80	008.08	0
008	Fiber, Acid Detergent (%)	0596	7.750	0.1000	8.175	0.5355	0.2530	61	-0.79	008.08	0
008	Fiber, Acid Detergent (%)	0939	7.820	0.4400	8.175	0.5355	0.2530	61	-0.66	008.08	0
008	Fiber, Acid Detergent (%)	0049	7.850	0.3400	8.175	0.5355	0.2530	61	-0.61	008.08	0
008	Fiber, Acid Detergent (%)	0294	7.850	0.1000	8.175	0.5355	0.2530	61	-0.61	008.08	0
008	Fiber, Acid Detergent (%)	0354	7.855	0.3500	8.175	0.5355	0.2530	61	-0.60	008.08	0
008	Fiber, Acid Detergent (%)	2196	7.890	0.0000	8.175	0.5355	0.2530	61	-0.53	008.08	0
008	Fiber, Acid Detergent (%)	0366	7.900	0.2000	8.175	0.5355	0.2530	61	-0.51	008.08	0
008	Fiber, Acid Detergent (%)	0278	7.950	0.3000	8.175	0.5355	0.2530	61	-0.42	008.08	0
008	Fiber, Acid Detergent (%)	0848	7.950	0.0400	8.175	0.5355	0.2530	61	-0.42	008.08	0
008	Fiber, Acid Detergent (%)	0581	7.950	0.2000	8.175	0.5355	0.2530	61	-0.42	008.08	0
008	Fiber, Acid Detergent (%)	0910	8.045	0.3900	8.175	0.5355	0.2530	61	-0.24	008.08	0
008	Fiber, Acid Detergent (%)	0510	8.050	0.1000	8.175	0.5355	0.2530	61	-0.23	008.08	0
008	Fiber, Acid Detergent (%)	0629	8.050	0.5000	8.175	0.5355	0.2530	61	-0.23	008.99	0
008	Fiber, Acid Detergent (%)	0353	8.070	0.1400	8.175	0.5355	0.2530	61	-0.20	008.02	0
008	Fiber, Acid Detergent (%)	0674	8.070	0.0000	8.175	0.5355	0.2530	61	-0.20	008.02	0
008	Fiber, Acid Detergent (%)	0004	8.070	0.2400	8.175	0.5355	0.2530	61	-0.20	008.08	0
008	Fiber, Acid Detergent (%)	0226	8.160	0.0400	8.175	0.5355	0.2530	61	-0.03	008.02	0
008	Fiber, Acid Detergent (%)	2190	8.160	0.1800	8.175	0.5355	0.2530	61	-0.03	008.08	0
008	Fiber, Acid Detergent (%)	0512	8.185	0.1720	8.175	0.5355	0.2530	61	0.02	008.02	0
008	Fiber, Acid Detergent (%)	2192	8.200	0.4000	8.175	0.5355	0.2530	61	0.05	008.02	0
008	Fiber, Acid Detergent (%)	0918	8.205	0.1300	8.175	0.5355	0.2530	61	0.06	008.08	0
008	Fiber, Acid Detergent (%)	0309	8.226	0.0142	8.175	0.5355	0.2530	61	0.09	008.02	0
008	Fiber, Acid Detergent (%)	0148	8.235	0.4700	8.175	0.5355	0.2530	61	0.11	008.08	0
008	Fiber, Acid Detergent (%)	0042	8.310	0.3800	8.175	0.5355	0.2530	61	0.25	008.08	0
008	Fiber, Acid Detergent (%)	0693	8.325	1.230	8.175	0.5355	0.2530	61	0.28	008.08	0
008	Fiber, Acid Detergent (%)	0689	8.350	0.7000	8.175	0.5355	0.2530	61	0.33	008.02	0
008	Fiber, Acid Detergent (%)	0948	8.355	0.1500	8.175	0.5355	0.2530	61	0.34	008.02	0
008	Fiber, Acid Detergent (%)	2246	8.355	0.1100	8.175	0.5355	0.2530	61	0.34	008.02	0
008	Fiber, Acid Detergent (%)	0265	8.370	0.5400	8.175	0.5355	0.2530	61	0.36	008.99	0
008	Fiber, Acid Detergent (%)	0083	8.375	0.0500	8.175	0.5355	0.2530	61	0.37	008.08	0
008	Fiber, Acid Detergent (%)	0035	8.385	0.0300	8.175	0.5355	0.2530	61	0.39	008.08	0
008	Fiber, Acid Detergent (%)	0504	8.415	0.6500	8.175	0.5355	0.2530	61	0.45	008.08	0
008	Fiber, Acid Detergent (%)	0171	8.500	0.0000	8.175	0.5355	0.2530	61	0.61	008.02	0
008	Fiber, Acid Detergent (%)	0297	8.500	0.6000	8.175	0.5355	0.2530	61	0.61	008.08	0
008	Fiber, Acid Detergent (%)	0001	8.512	0.2670	8.175	0.5355	0.2530	61	0.63	008.08	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
008	Fiber, Acid Detergent (%)	0357	8.560	0.0600	8.175	0.5355	0.2530	61	0.72	008.08	0
008	Fiber, Acid Detergent (%)	0263	8.588	0.0260	8.175	0.5355	0.2530	61	0.77	008.08	0
008	Fiber, Acid Detergent (%)	0066	8.655	0.2300	8.175	0.5355	0.2530	61	0.90	008.08	0
008	Fiber, Acid Detergent (%)	0190	8.770	0.0200	8.175	0.5355	0.2530	61	1.11	008.08	0
008	Fiber, Acid Detergent (%)	0726	8.770	0.0200	8.175	0.5355	0.2530	61	1.11	008.08	0
008	Fiber, Acid Detergent (%)	0870	8.809	0.5611	8.175	0.5355	0.2530	61	1.18	008.08	0
008	Fiber, Acid Detergent (%)	0148	8.835	0.2100	8.175	0.5355	0.2530	61	1.23	008.02	0
008	Fiber, Acid Detergent (%)	0912	8.840	0.1800	8.175	0.5355	0.2530	61	1.24	008.08	0
008	Fiber, Acid Detergent (%)	0610	8.860	0.5200	8.175	0.5355	0.2530	61	1.28	008.99	0
008	Fiber, Acid Detergent (%)	0511	8.900	0.6000	8.175	0.5355	0.2530	61	1.35	008.08	0
008	Fiber, Acid Detergent (%)	0407	8.917	0.0249	8.175	0.5355	0.2530	61	1.38	008.08	0
008	Fiber, Acid Detergent (%)	0164	9.050	0.3000	8.175	0.5355	0.2530	61	1.63	008.08	0
008	Fiber, Acid Detergent (%)	0045	9.300	0.7000	8.175	0.5355	0.2530	61	2.10	008.08	0
008	Fiber, Acid Detergent (%)	2290	9.400	0.2000	8.175	0.5355	0.2530	61	2.29	008.99	0
008	Fiber, Acid Detergent (%)	2375	13.10	10.04	8.175	0.5355	0.2530	61	9.20	008.08	2
009	Fiber, Neutral Detergent (%)	0598	17.49	0.0100	19.22	0.9892	0.4110	58	-1.75	009.09	0
009	Fiber, Neutral Detergent (%)	2129	17.81	0.3260	19.22	0.9892	0.4110	58	-1.42	009.09	0
009	Fiber, Neutral Detergent (%)	2375	18.05	0.1400	19.22	0.9892	0.4110	58	-1.18	009.09	0
009	Fiber, Neutral Detergent (%)	0596	18.15	0.3000	19.22	0.9892	0.4110	58	-1.08	009.09	0
009	Fiber, Neutral Detergent (%)	0512	18.16	1.110	19.22	0.9892	0.4110	58	-1.07	009.07	0
009	Fiber, Neutral Detergent (%)	0049	18.22	0.0800	19.22	0.9892	0.4110	58	-1.01	009.09	0
009	Fiber, Neutral Detergent (%)	2302	18.22	0.0000	19.22	0.9892	0.4110	58	-1.01	009.09	0
009	Fiber, Neutral Detergent (%)	0910	18.25	0.1000	19.22	0.9892	0.4110	58	-0.98	009.07	0
009	Fiber, Neutral Detergent (%)	0510	18.25	0.3000	19.22	0.9892	0.4110	58	-0.98	009.09	0
009	Fiber, Neutral Detergent (%)	0723	18.27	0.0900	19.22	0.9892	0.4110	58	-0.96	009.09	0
009	Fiber, Neutral Detergent (%)	0265	18.35	0.7000	19.22	0.9892	0.4110	58	-0.88	009.09	0
009	Fiber, Neutral Detergent (%)	0263	18.39	0.0040	19.22	0.9892	0.4110	58	-0.84	009.09	0
009	Fiber, Neutral Detergent (%)	0357	18.39	0.0000	19.22	0.9892	0.4110	58	-0.84	009.09	0
009	Fiber, Neutral Detergent (%)	2190	18.39	0.7800	19.22	0.9892	0.4110	58	-0.84	009.09	0
009	Fiber, Neutral Detergent (%)	0885	18.41	0.1404	19.22	0.9892	0.4110	58	-0.82	009.09	0
009	Fiber, Neutral Detergent (%)	0358	18.45	0.9900	19.22	0.9892	0.4110	58	-0.78	009.09	0
009	Fiber, Neutral Detergent (%)	0098	18.50	0.2000	19.22	0.9892	0.4110	58	-0.73	009.07	0
009	Fiber, Neutral Detergent (%)	2196	18.52	0.0000	19.22	0.9892	0.4110	58	-0.71	009.09	0
009	Fiber, Neutral Detergent (%)	0083	18.62	0.1500	19.22	0.9892	0.4110	58	-0.61	009.09	0
009	Fiber, Neutral Detergent (%)	0912	18.65	0.5600	19.22	0.9892	0.4110	58	-0.57	009.09	0
009	Fiber, Neutral Detergent (%)	0968	18.66	0.1000	19.22	0.9892	0.4110	58	-0.56	009.09	0
009	Fiber, Neutral Detergent (%)	0148	18.67	0.1600	19.22	0.9892	0.4110	58	-0.55	009.09	0
009	Fiber, Neutral Detergent (%)	0035	18.86	0.0800	19.22	0.9892	0.4110	58	-0.36	009.09	0
009	Fiber, Neutral Detergent (%)	0870	18.90	1.111	19.22	0.9892	0.4110	58	-0.33	009.09	0
009	Fiber, Neutral Detergent (%)	0294	18.95	0.3000	19.22	0.9892	0.4110	58	-0.27	009.09	0
009	Fiber, Neutral Detergent (%)	0345	19.00	0.4000	19.22	0.9892	0.4110	58	-0.22	009.07	0
009	Fiber, Neutral Detergent (%)	2022	19.04	0.2600	19.22	0.9892	0.4110	58	-0.18	009.09	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
009	Fiber, Neutral Detergent (%)	0956	19.10	0.0000	19.22	0.9892	0.4110	58	-0.12	009.09	0
009	Fiber, Neutral Detergent (%)	0504	19.24	0.1800	19.22	0.9892	0.4110	58	0.02	009.09	0
009	Fiber, Neutral Detergent (%)	0278	19.25	1.900	19.22	0.9892	0.4110	58	0.03	009.09	0
009	Fiber, Neutral Detergent (%)	0675	19.25	0.1000	19.22	0.9892	0.4110	58	0.03	009.09	0
009	Fiber, Neutral Detergent (%)	0918	19.27	0.7300	19.22	0.9892	0.4110	58	0.05	009.09	0
009	Fiber, Neutral Detergent (%)	0939	19.32	0.3900	19.22	0.9892	0.4110	58	0.10	009.09	0
009	Fiber, Neutral Detergent (%)	0353	19.38	0.3600	19.22	0.9892	0.4110	58	0.16	009.07	0
009	Fiber, Neutral Detergent (%)	0366	19.40	0.4000	19.22	0.9892	0.4110	58	0.18	009.09	0
009	Fiber, Neutral Detergent (%)	0629	19.40	0.4000	19.22	0.9892	0.4110	58	0.18	009.99	0
009	Fiber, Neutral Detergent (%)	0045	19.45	0.5000	19.22	0.9892	0.4110	58	0.23	009.09	0
009	Fiber, Neutral Detergent (%)	0164	19.45	0.3000	19.22	0.9892	0.4110	58	0.23	009.09	0
009	Fiber, Neutral Detergent (%)	0066	19.46	0.3700	19.22	0.9892	0.4110	58	0.24	009.09	0
009	Fiber, Neutral Detergent (%)	0015	19.58	0.3000	19.22	0.9892	0.4110	58	0.37	009.09	0
009	Fiber, Neutral Detergent (%)	0676	19.61	1.070	19.22	0.9892	0.4110	58	0.39	009.07	0
009	Fiber, Neutral Detergent (%)	0693	19.72	0.5000	19.22	0.9892	0.4110	58	0.51	009.09	0
009	Fiber, Neutral Detergent (%)	0581	19.79	1.220	19.22	0.9892	0.4110	58	0.58	009.09	0
009	Fiber, Neutral Detergent (%)	2192	19.80	0.2000	19.22	0.9892	0.4110	58	0.59	009.07	0
009	Fiber, Neutral Detergent (%)	0610	20.05	0.1000	19.22	0.9892	0.4110	58	0.84	009.07	0
009	Fiber, Neutral Detergent (%)	0848	20.06	0.1900	19.22	0.9892	0.4110	58	0.85	009.09	0
009	Fiber, Neutral Detergent (%)	0190	20.08	0.0500	19.22	0.9892	0.4110	58	0.87	009.09	0
009	Fiber, Neutral Detergent (%)	0726	20.08	0.0500	19.22	0.9892	0.4110	58	0.87	009.09	0
009	Fiber, Neutral Detergent (%)	0674	20.25	0.0000	19.22	0.9892	0.4110	58	1.04	009.07	0
009	Fiber, Neutral Detergent (%)	0354	20.39	0.7400	19.22	0.9892	0.4110	58	1.18	009.09	0
009	Fiber, Neutral Detergent (%)	0948	20.57	0.6800	19.22	0.9892	0.4110	58	1.37	009.07	0
009	Fiber, Neutral Detergent (%)	0226	20.74	0.4500	19.22	0.9892	0.4110	58	1.53	009.07	0
009	Fiber, Neutral Detergent (%)	0297	21.00	2.000	19.22	0.9892	0.4110	58	1.80	009.09	0
009	Fiber, Neutral Detergent (%)	0689	21.10	0.6000	19.22	0.9892	0.4110	58	1.90	009.07	0
009	Fiber, Neutral Detergent (%)	0407	21.33	0.5159	19.22	0.9892	0.4110	58	2.14	009.09	0
009	Fiber, Neutral Detergent (%)	0309	21.86	0.1590	19.22	0.9892	0.4110	58	2.67	009.07	0
009	Fiber, Neutral Detergent (%)	2290	22.10	0.2000	19.22	0.9892	0.4110	58	2.91	009.99	0
009	Fiber, Neutral Detergent (%)	2246	22.31	0.7900	19.22	0.9892	0.4110	58	3.12	009.07	0
009	Fiber, Neutral Detergent (%)	0638	14.65	9.300	19.22	0.9892	0.4110	58	-4.62	009.09	1
010	Moisture (%)	2129	8.155	0.0010	9.499	0.3690	0.0934	15	-3.64	010.99	0
010	Moisture (%)	0019	8.520	0.4200	9.499	0.3690	0.0934	15	-2.65	010.99	0
010	Moisture (%)	2290	9.290	0.1000	9.499	0.3690	0.0934	15	-0.57	010.11	0
010	Moisture (%)	0939	9.315	0.1700	9.499	0.3690	0.0934	15	-0.50	010.99	0
010	Moisture (%)	0726	9.340	0.0400	9.499	0.3690	0.0934	15	-0.43	010.99	0
010	Moisture (%)	0843	9.385	0.0300	9.499	0.3690	0.0934	15	-0.31	010.03	0
010	Moisture (%)	0948	9.440	0.0400	9.499	0.3690	0.0934	15	-0.16	010.99	0
010	Moisture (%)	0885	9.445	0.0100	9.499	0.3690	0.0934	15	-0.15	010.11	0
010	Moisture (%)	2076	9.570	0.0000	9.499	0.3690	0.0934	15	0.19	010.99	0
010	Moisture (%)	2150	9.605	0.1700	9.499	0.3690	0.0934	15	0.29	010.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
010	Moisture (%)	2375	9.605	0.0100	9.499	0.3690	0.0934	15	0.29	010.99	0
010	Moisture (%)	0652	9.700	0.2000	9.499	0.3690	0.0934	15	0.54	010.99	0
010	Moisture (%)	0164	9.795	0.0700	9.499	0.3690	0.0934	15	0.80	010.03	0
010	Moisture (%)	0190	10.06	0.0000	9.499	0.3690	0.0934	15	1.52	010.99	0
010	Moisture (%)	2089	10.13	0.1400	9.499	0.3690	0.0934	15	1.71	010.99	0
010	Moisture (%)	0692	9.050	0.9000	9.499	0.3690	0.0934	15	-1.22	010.99	1
011	Loss on Drying, HT (%)	2359	8.500	0.2000	10.71	0.5253	0.1480	67	-4.20	011.01	0
011	Loss on Drying, HT (%)	2314	8.675	0.0100	10.71	0.5253	0.1480	67	-3.87	011.01	0
011	Loss on Drying, HT (%)	0618	9.495	0.0700	10.71	0.5253	0.1480	67	-2.31	011.01	0
011	Loss on Drying, HT (%)	2200	9.500	0.4000	10.71	0.5253	0.1480	67	-2.30	011.01	0
011	Loss on Drying, HT (%)	0859	9.613	0.0690	10.71	0.5253	0.1480	67	-2.09	011.01	0
011	Loss on Drying, HT (%)	0047	9.850	0.1000	10.71	0.5253	0.1480	67	-1.63	011.01	0
011	Loss on Drying, HT (%)	2199	9.950	0.7000	10.71	0.5253	0.1480	67	-1.44	011.01	0
011	Loss on Drying, HT (%)	0520	10.09	0.0300	10.71	0.5253	0.1480	67	-1.19	011.01	0
011	Loss on Drying, HT (%)	0843	10.11	0.3700	10.71	0.5253	0.1480	67	-1.15	011.01	0
011	Loss on Drying, HT (%)	0643	10.18	0.0740	10.71	0.5253	0.1480	67	-1.01	011.01	0
011	Loss on Drying, HT (%)	0942	10.21	0.0500	10.71	0.5253	0.1480	67	-0.96	011.02	0
011	Loss on Drying, HT (%)	2340	10.28	0.0300	10.71	0.5253	0.1480	67	-0.82	011.01	0
011	Loss on Drying, HT (%)	2303	10.29	0.3200	10.71	0.5253	0.1480	67	-0.80	011.99	0
011	Loss on Drying, HT (%)	0848	10.32	0.0500	10.71	0.5253	0.1480	67	-0.75	011.01	0
011	Loss on Drying, HT (%)	0968	10.32	0.0100	10.71	0.5253	0.1480	67	-0.75	011.01	0
011	Loss on Drying, HT (%)	2326	10.32	0.4800	10.71	0.5253	0.1480	67	-0.74	011.01	0
011	Loss on Drying, HT (%)	0953	10.36	0.2400	10.71	0.5253	0.1480	67	-0.66	011.01	0
011	Loss on Drying, HT (%)	0536	10.38	0.2100	10.71	0.5253	0.1480	67	-0.63	011.01	0
011	Loss on Drying, HT (%)	0354	10.41	0.0100	10.71	0.5253	0.1480	67	-0.58	011.01	0
011	Loss on Drying, HT (%)	2196	10.44	0.0000	10.71	0.5253	0.1480	67	-0.51	011.01	0
011	Loss on Drying, HT (%)	0171	10.45	0.0100	10.71	0.5253	0.1480	67	-0.50	011.01	0
011	Loss on Drying, HT (%)	0539	10.46	0.3300	10.71	0.5253	0.1480	67	-0.48	011.01	0
011	Loss on Drying, HT (%)	0674	10.47	0.0000	10.71	0.5253	0.1480	67	-0.45	011.01	0
011	Loss on Drying, HT (%)	0675	10.48	0.0100	10.71	0.5253	0.1480	67	-0.44	011.01	0
011	Loss on Drying, HT (%)	0574	10.51	0.0700	10.71	0.5253	0.1480	67	-0.39	011.01	0
011	Loss on Drying, HT (%)	0098	10.52	0.0300	10.71	0.5253	0.1480	67	-0.37	011.01	0
011	Loss on Drying, HT (%)	0723	10.54	0.0900	10.71	0.5253	0.1480	67	-0.33	011.01	0
011	Loss on Drying, HT (%)	0100	10.54	0.1600	10.71	0.5253	0.1480	67	-0.32	011.01	0
011	Loss on Drying, HT (%)	0229	10.63	1.050	10.71	0.5253	0.1480	67	-0.16	011.01	0
011	Loss on Drying, HT (%)	0529	10.64	0.3300	10.71	0.5253	0.1480	67	-0.14	011.02	0
011	Loss on Drying, HT (%)	0263	10.66	0.0050	10.71	0.5253	0.1480	67	-0.09	011.01	0
011	Loss on Drying, HT (%)	2337	10.67	0.0200	10.71	0.5253	0.1480	67	-0.07	011.01	0
011	Loss on Drying, HT (%)	2188	10.71	0.0000	10.71	0.5253	0.1480	67	0.00	011.01	0
011	Loss on Drying, HT (%)	0425	10.73	0.0200	10.71	0.5253	0.1480	67	0.04	011.01	0
011	Loss on Drying, HT (%)	0598	10.76	0.0800	10.71	0.5253	0.1480	67	0.10	011.01	0
011	Loss on Drying, HT (%)	0006	10.81	0.3990	10.71	0.5253	0.1480	67	0.19	011.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
011	Loss on Drying, HT (%)	2268	10.81	0.3000	10.71	0.5253	0.1480	67	0.19	011.01	0
011	Loss on Drying, HT (%)	0553	10.82	0.5800	10.71	0.5253	0.1480	67	0.21	011.01	0
011	Loss on Drying, HT (%)	0358	10.83	0.4300	10.71	0.5253	0.1480	67	0.22	011.01	0
011	Loss on Drying, HT (%)	0294	10.85	0.1000	10.71	0.5253	0.1480	67	0.27	011.01	0
011	Loss on Drying, HT (%)	0511	10.88	0.1700	10.71	0.5253	0.1480	67	0.32	011.01	0
011	Loss on Drying, HT (%)	0062	10.90	0.0200	10.71	0.5253	0.1480	67	0.37	011.01	0
011	Loss on Drying, HT (%)	2246	10.91	0.1700	10.71	0.5253	0.1480	67	0.37	011.01	0
011	Loss on Drying, HT (%)	0066	10.91	0.1600	10.71	0.5253	0.1480	67	0.38	011.01	0
011	Loss on Drying, HT (%)	0870	10.91	0.0305	10.71	0.5253	0.1480	67	0.39	011.01	0
011	Loss on Drying, HT (%)	2006	10.93	0.0800	10.71	0.5253	0.1480	67	0.42	011.01	0
011	Loss on Drying, HT (%)	0651	10.97	0.0602	10.71	0.5253	0.1480	67	0.49	011.01	0
011	Loss on Drying, HT (%)	0226	10.99	0.0500	10.71	0.5253	0.1480	67	0.53	011.01	0
011	Loss on Drying, HT (%)	2184	11.04	0.0100	10.71	0.5253	0.1480	67	0.62	011.01	0
011	Loss on Drying, HT (%)	2397	11.04	0.0200	10.71	0.5253	0.1480	67	0.63	011.01	0
011	Loss on Drying, HT (%)	2193	11.07	0.0300	10.71	0.5253	0.1480	67	0.68	011.01	0
011	Loss on Drying, HT (%)	0242	11.08	0.0600	10.71	0.5253	0.1480	67	0.71	011.01	0
011	Loss on Drying, HT (%)	0164	11.09	0.0200	10.71	0.5253	0.1480	67	0.73	011.01	0
011	Loss on Drying, HT (%)	2302	11.12	0.0200	10.71	0.5253	0.1480	67	0.78	011.01	0
011	Loss on Drying, HT (%)	0510	11.15	0.1000	10.71	0.5253	0.1480	67	0.84	011.01	0
011	Loss on Drying, HT (%)	2073	11.17	0.0200	10.71	0.5253	0.1480	67	0.88	011.01	0
011	Loss on Drying, HT (%)	0541	11.18	0.0200	10.71	0.5253	0.1480	67	0.90	011.01	0
011	Loss on Drying, HT (%)	0123	11.20	0.0000	10.71	0.5253	0.1480	67	0.94	011.01	0
011	Loss on Drying, HT (%)	0309	11.32	0.2900	10.71	0.5253	0.1480	67	1.16	011.01	0
011	Loss on Drying, HT (%)	0589	11.35	0.1100	10.71	0.5253	0.1480	67	1.21	011.01	0
011	Loss on Drying, HT (%)	0202	11.39	0.1200	10.71	0.5253	0.1480	67	1.30	011.01	0
011	Loss on Drying, HT (%)	0144	11.42	0.0700	10.71	0.5253	0.1480	67	1.35	011.01	0
011	Loss on Drying, HT (%)	0108	11.57	0.0400	10.71	0.5253	0.1480	67	1.64	011.01	0
011	Loss on Drying, HT (%)	0265	11.60	0.4000	10.71	0.5253	0.1480	67	1.70	011.01	0
011	Loss on Drying, HT (%)	0004	11.76	0.0900	10.71	0.5253	0.1480	67	1.99	011.01	0
011	Loss on Drying, HT (%)	0407	12.39	0.3109	10.71	0.5253	0.1480	67	3.19	011.01	0
011	Loss on Drying, HT (%)	2022	12.52	0.0100	10.71	0.5253	0.1480	67	3.44	011.01	0
011	Loss on Drying, HT (%)	2174	10.58	1.490	10.71	0.5253	0.1480	67	-0.25	011.99	1
011	Loss on Drying, HT (%)	2174	11.03	1.190	10.71	0.5253	0.1480	67	0.60	011.01	1
011	Loss on Drying, HT (%)	2168	11.26	1.110	10.71	0.5253	0.1480	67	1.04	011.99	1
012	Starch (%)	0265	16.45	0.7000	29.34	1.490	0.5433	40	-8.65	012.01	0
012	Starch (%)	0676	26.48	1.660	29.34	1.490	0.5433	40	-1.92	012.01	0
012	Starch (%)	0918	26.76	1.908	29.34	1.490	0.5433	40	-1.73	012.01	0
012	Starch (%)	0066	27.37	0.4900	29.34	1.490	0.5433	40	-1.32	012.03	0
012	Starch (%)	0407	27.37	1.141	29.34	1.490	0.5433	40	-1.32	012.03	0
012	Starch (%)	0956	27.70	0.6000	29.34	1.490	0.5433	40	-1.10	012.01	0
012	Starch (%)	0870	28.19	0.5632	29.34	1.490	0.5433	40	-0.77	012.01	0
012	Starch (%)	0948	28.26	0.5900	29.34	1.490	0.5433	40	-0.73	012.04	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
012	Starch (%)	2129	28.32	0.3790	29.34	1.490	0.5433	40	-0.68	012.04	0
012	Starch (%)	2188	28.40	0.0000	29.34	1.490	0.5433	40	-0.63	012.00	0
012	Starch (%)	0297	28.40	0.7400	29.34	1.490	0.5433	40	-0.63	012.03	0
012	Starch (%)	0510	28.40	0.6000	29.34	1.490	0.5433	40	-0.63	012.04	0
012	Starch (%)	2290	28.50	0.4000	29.34	1.490	0.5433	40	-0.56	012.11	0
012	Starch (%)	0353	28.62	0.0200	29.34	1.490	0.5433	40	-0.48	012.20	0
012	Starch (%)	0226	28.67	0.5200	29.34	1.490	0.5433	40	-0.45	012.04	0
012	Starch (%)	0164	28.95	1.300	29.34	1.490	0.5433	40	-0.26	012.01	0
012	Starch (%)	0164	29.05	1.100	29.34	1.490	0.5433	40	-0.19	012.00	0
012	Starch (%)	2193	29.14	0.1100	29.34	1.490	0.5433	40	-0.14	012.00	0
012	Starch (%)	0723	29.17	0.1600	29.34	1.490	0.5433	40	-0.11	012.00	0
012	Starch (%)	0848	29.26	0.0900	29.34	1.490	0.5433	40	-0.06	012.00	0
012	Starch (%)	0045	29.35	0.3000	29.34	1.490	0.5433	40	0.01	012.01	0
012	Starch (%)	0278	29.60	0.0000	29.34	1.490	0.5433	40	0.18	012.04	0
012	Starch (%)	0910	29.65	1.900	29.34	1.490	0.5433	40	0.21	012.00	0
012	Starch (%)	0354	29.74	0.0270	29.34	1.490	0.5433	40	0.27	012.00	0
012	Starch (%)	0148	29.76	1.560	29.34	1.490	0.5433	40	0.28	012.01	0
012	Starch (%)	2319	29.80	0.6000	29.34	1.490	0.5433	40	0.31	012.99	0
012	Starch (%)	2192	29.99	0.5500	29.34	1.490	0.5433	40	0.43	012.00	0
012	Starch (%)	0098	30.00	0.4000	29.34	1.490	0.5433	40	0.44	012.04	0
012	Starch (%)	0689	30.15	0.1000	29.34	1.490	0.5433	40	0.54	012.00	0
012	Starch (%)	0610	30.25	0.7000	29.34	1.490	0.5433	40	0.61	012.00	0
012	Starch (%)	0683	30.33	0.0000	29.34	1.490	0.5433	40	0.67	012.00	0
012	Starch (%)	0619	30.35	0.1000	29.34	1.490	0.5433	40	0.68	012.00	0
012	Starch (%)	0674	30.51	0.0000	29.34	1.490	0.5433	40	0.79	012.04	0
012	Starch (%)	0693	30.61	0.1500	29.34	1.490	0.5433	40	0.85	012.01	0
012	Starch (%)	2009	31.13	0.0394	29.34	1.490	0.5433	40	1.20	012.00	0
012	Starch (%)	0407	31.16	0.4212	29.34	1.490	0.5433	40	1.22	012.01	0
012	Starch (%)	2006	31.40	0.8900	29.34	1.490	0.5433	40	1.38	012.00	0
012	Starch (%)	2089	32.60	0.2500	29.34	1.490	0.5433	40	2.19	012.01	0
012	Starch (%)	0297	33.23	0.1900	29.34	1.490	0.5433	40	2.61	012.11	0
012	Starch (%)	0885	33.61	0.4842	29.34	1.490	0.5433	40	2.87	012.01	0
012	Starch (%)	0638	24.60	9.200	29.34	1.490	0.5433	40	-3.18	012.00	1
013	Fat, Acid Pretreat (%)	0618	2.135	0.0900	4.546	0.6688	0.2318	48	-3.60	013.08	0
013	Fat, Acid Pretreat (%)	2314	3.145	0.0100	4.546	0.6688	0.2318	48	-2.09	013.00	0
013	Fat, Acid Pretreat (%)	2073	3.190	0.0400	4.546	0.6688	0.2318	48	-2.03	013.00	0
013	Fat, Acid Pretreat (%)	0265	3.205	0.0300	4.546	0.6688	0.2318	48	-2.00	013.13	0
013	Fat, Acid Pretreat (%)	0353	3.450	2.060	4.546	0.6688	0.2318	48	-1.64	013.10	0
013	Fat, Acid Pretreat (%)	0914	3.730	0.6800	4.546	0.6688	0.2318	48	-1.22	013.02	0
013	Fat, Acid Pretreat (%)	0689	3.950	0.3000	4.546	0.6688	0.2318	48	-0.89	013.00	0
013	Fat, Acid Pretreat (%)	2268	3.950	0.3000	4.546	0.6688	0.2318	48	-0.89	013.10	0
013	Fat, Acid Pretreat (%)	0675	4.045	0.0100	4.546	0.6688	0.2318	48	-0.75	013.13	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
013	Fat, Acid Pretreat (%)	2196	4.100	0.0000	4.546	0.6688	0.2318	48	-0.67	013.10	0
013	Fat, Acid Pretreat (%)	0083	4.110	0.1600	4.546	0.6688	0.2318	48	-0.65	013.00	0
013	Fat, Acid Pretreat (%)	0036	4.161	0.0183	4.546	0.6688	0.2318	48	-0.58	013.10	0
013	Fat, Acid Pretreat (%)	0581	4.175	0.2700	4.546	0.6688	0.2318	48	-0.55	013.13	0
013	Fat, Acid Pretreat (%)	2319	4.250	0.1000	4.546	0.6688	0.2318	48	-0.44	013.00	0
013	Fat, Acid Pretreat (%)	0242	4.260	0.0400	4.546	0.6688	0.2318	48	-0.43	013.00	0
013	Fat, Acid Pretreat (%)	0939	4.260	0.3400	4.546	0.6688	0.2318	48	-0.43	013.13	0
013	Fat, Acid Pretreat (%)	0610	4.270	0.0200	4.546	0.6688	0.2318	48	-0.41	013.10	0
013	Fat, Acid Pretreat (%)	2146	4.405	0.0700	4.546	0.6688	0.2318	48	-0.21	013.00	0
013	Fat, Acid Pretreat (%)	0354	4.405	0.1700	4.546	0.6688	0.2318	48	-0.21	013.13	0
013	Fat, Acid Pretreat (%)	0297	4.410	0.0200	4.546	0.6688	0.2318	48	-0.20	013.00	0
013	Fat, Acid Pretreat (%)	0650	4.410	0.1200	4.546	0.6688	0.2318	48	-0.20	013.00	0
013	Fat, Acid Pretreat (%)	2129	4.456	0.2280	4.546	0.6688	0.2318	48	-0.13	013.13	0
013	Fat, Acid Pretreat (%)	0968	4.490	0.0000	4.546	0.6688	0.2318	48	-0.08	013.13	0
013	Fat, Acid Pretreat (%)	2076	4.497	0.0018	4.546	0.6688	0.2318	48	-0.07	013.00	0
013	Fat, Acid Pretreat (%)	0300	4.535	0.2100	4.546	0.6688	0.2318	48	-0.02	013.13	0
013	Fat, Acid Pretreat (%)	0652	4.550	0.1000	4.546	0.6688	0.2318	48	0.01	013.00	0
013	Fat, Acid Pretreat (%)	0910	4.575	0.0900	4.546	0.6688	0.2318	48	0.04	013.00	0
013	Fat, Acid Pretreat (%)	0948	4.590	0.0200	4.546	0.6688	0.2318	48	0.07	013.00	0
013	Fat, Acid Pretreat (%)	0683	4.590	0.1800	4.546	0.6688	0.2318	48	0.07	013.13	0
013	Fat, Acid Pretreat (%)	0229	4.595	0.1500	4.546	0.6688	0.2318	48	0.07	013.00	0
013	Fat, Acid Pretreat (%)	0843	4.640	0.9000	4.546	0.6688	0.2318	48	0.14	013.13	0
013	Fat, Acid Pretreat (%)	0098	4.695	0.5700	4.546	0.6688	0.2318	48	0.22	013.02	0
013	Fat, Acid Pretreat (%)	0638	4.700	0.6000	4.546	0.6688	0.2318	48	0.23	013.10	0
013	Fat, Acid Pretreat (%)	0541	4.780	0.1800	4.546	0.6688	0.2318	48	0.35	013.00	0
013	Fat, Acid Pretreat (%)	0148	4.915	1.290	4.546	0.6688	0.2318	48	0.55	013.02	0
013	Fat, Acid Pretreat (%)	0643	4.920	0.0200	4.546	0.6688	0.2318	48	0.56	013.02	0
013	Fat, Acid Pretreat (%)	0870	4.926	0.0363	4.546	0.6688	0.2318	48	0.57	013.02	0
013	Fat, Acid Pretreat (%)	0553	4.965	0.0700	4.546	0.6688	0.2318	48	0.63	013.02	0
013	Fat, Acid Pretreat (%)	0164	5.080	0.1200	4.546	0.6688	0.2318	48	0.80	013.02	0
013	Fat, Acid Pretreat (%)	2259	5.115	0.0010	4.546	0.6688	0.2318	48	0.85	013.02	0
013	Fat, Acid Pretreat (%)	0171	5.380	0.2600	4.546	0.6688	0.2318	48	1.25	013.02	0
013	Fat, Acid Pretreat (%)	0407	5.409	0.1200	4.546	0.6688	0.2318	48	1.29	013.02	0
013	Fat, Acid Pretreat (%)	0407	5.494	0.2349	4.546	0.6688	0.2318	48	1.42	013.13	0
013	Fat, Acid Pretreat (%)	0309	5.500	0.2538	4.546	0.6688	0.2318	48	1.43	013.00	0
013	Fat, Acid Pretreat (%)	0682	5.580	0.0000	4.546	0.6688	0.2318	48	1.55	013.02	0
013	Fat, Acid Pretreat (%)	0676	5.765	0.3500	4.546	0.6688	0.2318	48	1.82	013.02	0
013	Fat, Acid Pretreat (%)	0202	5.870	0.1600	4.546	0.6688	0.2318	48	1.98	013.00	0
013	Fat, Acid Pretreat (%)	2150	6.025	0.1300	4.546	0.6688	0.2318	48	2.21	013.00	0
013	Fat, Acid Pretreat (%)	0876	6.781	4.762	4.546	0.6688	0.2318	48	3.34	013.13	1
015	Aluminum (ppm)	0123	51.60	0.0000	85.69	21.83	1.786	14	-1.56	015.99	0
015	Aluminum (ppm)	0510	59.00	2.000	85.69	21.83	1.786	14	-1.22	015.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
015	Aluminum (ppm)	0049	61.50	0.3400	85.69	21.83	1.786	14	-1.11	015.41	0
015	Aluminum (ppm)	0520	64.29	2.021	85.69	21.83	1.786	14	-0.98	015.41	0
015	Aluminum (ppm)	2129	84.94	1.620	85.69	21.83	1.786	14	-0.03	015.42	0
015	Aluminum (ppm)	0407	88.65	3.014	85.69	21.83	1.786	14	0.14	015.41	0
015	Aluminum (ppm)	0345	90.68	0.6500	85.69	21.83	1.786	14	0.23	015.43	0
015	Aluminum (ppm)	0407	91.60	1.716	85.69	21.83	1.786	14	0.27	015.43	0
015	Aluminum (ppm)	0171	91.75	0.4200	85.69	21.83	1.786	14	0.28	015.41	0
015	Aluminum (ppm)	0164	93.00	2.000	85.69	21.83	1.786	14	0.33	015.41	0
015	Aluminum (ppm)	0169	96.15	1.100	85.69	21.83	1.786	14	0.48	015.43	0
015	Aluminum (ppm)	0407	99.69	2.127	85.69	21.83	1.786	14	0.64	015.53	0
015	Aluminum (ppm)	0553	107.0	2.000	85.69	21.83	1.786	14	0.98	015.53	0
015	Aluminum (ppm)	0353	125.0	6.000	85.69	21.83	1.786	14	1.80	015.43	0
015	Aluminum (ppm)	0297	83.50	15.00	85.69	21.83	1.786	14	-0.10	015.43	1
017	Boron (ppm)	0226	6.530	0.0000	7.572	0.8145	0.4365	16	-1.28	017.41	0
017	Boron (ppm)	0345	6.675	0.0500	7.572	0.8145	0.4365	16	-1.10	017.43	0
017	Boron (ppm)	0693	6.750	0.3800	7.572	0.8145	0.4365	16	-1.01	017.42	0
017	Boron (ppm)	0510	7.000	0.0000	7.572	0.8145	0.4365	16	-0.70	017.43	0
017	Boron (ppm)	0407	7.208	0.3114	7.572	0.8145	0.4365	16	-0.45	017.41	0
017	Boron (ppm)	0358	7.275	0.1500	7.572	0.8145	0.4365	16	-0.36	017.41	0
017	Boron (ppm)	0049	7.290	1.520	7.572	0.8145	0.4365	16	-0.35	017.41	0
017	Boron (ppm)	0407	7.294	0.1875	7.572	0.8145	0.4365	16	-0.34	017.53	0
017	Boron (ppm)	0045	7.410	0.9000	7.572	0.8145	0.4365	16	-0.20	017.42	0
017	Boron (ppm)	0294	7.610	0.4200	7.572	0.8145	0.4365	16	0.05	017.42	0
017	Boron (ppm)	0265	7.900	0.0000	7.572	0.8145	0.4365	16	0.40	017.42	0
017	Boron (ppm)	0123	7.980	0.0000	7.572	0.8145	0.4365	16	0.50	017.99	0
017	Boron (ppm)	0407	8.139	0.8244	7.572	0.8145	0.4365	16	0.70	017.43	0
017	Boron (ppm)	0353	8.500	1.000	7.572	0.8145	0.4365	16	1.14	017.43	0
017	Boron (ppm)	0083	8.900	0.1000	7.572	0.8145	0.4365	16	1.63	017.43	0
017	Boron (ppm)	0229	10.03	1.140	7.572	0.8145	0.4365	16	3.02	017.41	0
017	Boron (ppm)	2129	7.337	2.625	7.572	0.8145	0.4365	16	-0.29	017.42	1
017	Boron (ppm)	0297	0.0000	0.0000	7.572	0.8145	0.4365	16		017.43	4
019	Calcium (%)	0226	0.9950	0.0100	1.184	0.0580	0.0244	122	-3.26	019.43	0
019	Calcium (%)	0018	1.035	0.0700	1.184	0.0580	0.0244	122	-2.57	019.43	0
019	Calcium (%)	2006	1.040	0.0390	1.184	0.0580	0.0244	122	-2.49	019.00	0
019	Calcium (%)	0169	1.055	0.0300	1.184	0.0580	0.0244	122	-2.22	019.32	0
019	Calcium (%)	0529	1.065	0.0100	1.184	0.0580	0.0244	122	-2.05	019.31	0
019	Calcium (%)	0242	1.065	0.0100	1.184	0.0580	0.0244	122	-2.05	019.99	0
019	Calcium (%)	2129	1.067	0.0320	1.184	0.0580	0.0244	122	-2.01	019.42	0
019	Calcium (%)	0047	1.070	0.0400	1.184	0.0580	0.0244	122	-1.96	019.52	0
019	Calcium (%)	0066	1.078	0.0420	1.184	0.0580	0.0244	122	-1.82	019.41	0
019	Calcium (%)	2006	1.091	0.0490	1.184	0.0580	0.0244	122	-1.61	019.09	0
019	Calcium (%)	2109	1.098	0.0480	1.184	0.0580	0.0244	122	-1.48	019.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
019	Calcium (%)	2188	1.100	0.0000	1.184	0.0580	0.0244	122	-1.45	019.08	0
019	Calcium (%)	0033	1.105	0.0100	1.184	0.0580	0.0244	122	-1.36	019.43	0
019	Calcium (%)	2192	1.105	0.0300	1.184	0.0580	0.0244	122	-1.36	019.43	0
019	Calcium (%)	0656	1.110	0.0000	1.184	0.0580	0.0244	122	-1.27	019.31	0
019	Calcium (%)	0263	1.113	0.0050	1.184	0.0580	0.0244	122	-1.23	019.42	0
019	Calcium (%)	2246	1.120	0.0000	1.184	0.0580	0.0244	122	-1.10	019.31	0
019	Calcium (%)	0229	1.120	0.0000	1.184	0.0580	0.0244	122	-1.10	019.41	0
019	Calcium (%)	0511	1.120	0.0600	1.184	0.0580	0.0244	122	-1.10	019.43	0
019	Calcium (%)	2325	1.130	0.0200	1.184	0.0580	0.0244	122	-0.93	019.31	0
019	Calcium (%)	0910	1.130	0.0400	1.184	0.0580	0.0244	122	-0.93	019.41	0
019	Calcium (%)	0357	1.133	0.0146	1.184	0.0580	0.0244	122	-0.88	019.42	0
019	Calcium (%)	0345	1.135	0.0100	1.184	0.0580	0.0244	122	-0.84	019.43	0
019	Calcium (%)	0675	1.135	0.0100	1.184	0.0580	0.0244	122	-0.84	019.43	0
019	Calcium (%)	0358	1.140	0.0000	1.184	0.0580	0.0244	122	-0.76	019.41	0
019	Calcium (%)	0619	1.140	0.0400	1.184	0.0580	0.0244	122	-0.76	019.41	0
019	Calcium (%)	0520	1.150	0.0180	1.184	0.0580	0.0244	122	-0.58	019.41	0
019	Calcium (%)	0006	1.150	0.0000	1.184	0.0580	0.0244	122	-0.58	019.43	0
019	Calcium (%)	0651	1.151	0.0563	1.184	0.0580	0.0244	122	-0.57	019.31	0
019	Calcium (%)	0512	1.151	0.0200	1.184	0.0580	0.0244	122	-0.57	019.43	0
019	Calcium (%)	2319	1.155	0.0100	1.184	0.0580	0.0244	122	-0.50	019.42	0
019	Calcium (%)	0870	1.156	0.0100	1.184	0.0580	0.0244	122	-0.48	019.42	0
019	Calcium (%)	0036	1.159	0.0401	1.184	0.0580	0.0244	122	-0.42	019.43	0
019	Calcium (%)	0629	1.160	0.0000	1.184	0.0580	0.0244	122	-0.41	019.08	0
019	Calcium (%)	0038	1.160	0.0400	1.184	0.0580	0.0244	122	-0.41	019.31	0
019	Calcium (%)	0948	1.160	0.0000	1.184	0.0580	0.0244	122	-0.41	019.31	0
019	Calcium (%)	0035	1.160	0.0200	1.184	0.0580	0.0244	122	-0.41	019.42	0
019	Calcium (%)	0968	1.163	0.0030	1.184	0.0580	0.0244	122	-0.37	019.43	0
019	Calcium (%)	2146	1.165	0.0100	1.184	0.0580	0.0244	122	-0.32	019.00	0
019	Calcium (%)	0004	1.165	0.0500	1.184	0.0580	0.0244	122	-0.32	019.41	0
019	Calcium (%)	0226	1.165	0.0300	1.184	0.0580	0.0244	122	-0.32	019.41	0
019	Calcium (%)	0294	1.165	0.0100	1.184	0.0580	0.0244	122	-0.32	019.42	0
019	Calcium (%)	0693	1.165	0.0500	1.184	0.0580	0.0244	122	-0.32	019.42	0
019	Calcium (%)	0010	1.165	0.0700	1.184	0.0580	0.0244	122	-0.32	019.43	0
019	Calcium (%)	0629	1.165	0.0300	1.184	0.0580	0.0244	122	-0.32	019.43	0
019	Calcium (%)	0676	1.165	0.0100	1.184	0.0580	0.0244	122	-0.32	019.99	0
019	Calcium (%)	0870	1.167	0.0640	1.184	0.0580	0.0244	122	-0.29	019.43	0
019	Calcium (%)	2196	1.170	0.0000	1.184	0.0580	0.0244	122	-0.24	019.08	0
019	Calcium (%)	0019	1.170	0.0000	1.184	0.0580	0.0244	122	-0.24	019.41	0
019	Calcium (%)	0144	1.170	0.0200	1.184	0.0580	0.0244	122	-0.24	019.41	0
019	Calcium (%)	0674	1.170	0.0000	1.184	0.0580	0.0244	122	-0.24	019.41	0
019	Calcium (%)	0202	1.170	0.0600	1.184	0.0580	0.0244	122	-0.24	019.43	0
019	Calcium (%)	0297	1.170	0.0400	1.184	0.0580	0.0244	122	-0.24	019.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
019	Calcium (%)	0407	1.172	0.0293	1.184	0.0580	0.0244	122	-0.21	019.43	0
019	Calcium (%)	2009	1.172	0.0109	1.184	0.0580	0.0244	122	-0.20	019.08	0
019	Calcium (%)	0001	1.175	0.0070	1.184	0.0580	0.0244	122	-0.16	019.31	0
019	Calcium (%)	0848	1.175	0.0100	1.184	0.0580	0.0244	122	-0.15	019.41	0
019	Calcium (%)	0083	1.175	0.0100	1.184	0.0580	0.0244	122	-0.15	019.43	0
019	Calcium (%)	0164	1.175	0.0100	1.184	0.0580	0.0244	122	-0.15	019.44	0
019	Calcium (%)	0123	1.175	0.0300	1.184	0.0580	0.0244	122	-0.15	019.99	0
019	Calcium (%)	2199	1.180	0.0400	1.184	0.0580	0.0244	122	-0.07	019.08	0
019	Calcium (%)	0089	1.180	0.0000	1.184	0.0580	0.0244	122	-0.07	019.43	0
019	Calcium (%)	2340	1.181	0.0080	1.184	0.0580	0.0244	122	-0.05	019.08	0
019	Calcium (%)	0407	1.184	0.0432	1.184	0.0580	0.0244	122	0.00	019.53	0
019	Calcium (%)	2146	1.185	0.0700	1.184	0.0580	0.0244	122	0.02	019.31	0
019	Calcium (%)	0035	1.185	0.0100	1.184	0.0580	0.0244	122	0.02	019.43	0
019	Calcium (%)	0876	1.185	0.0700	1.184	0.0580	0.0244	122	0.02	019.43	0
019	Calcium (%)	2340	1.189	0.0030	1.184	0.0580	0.0244	122	0.08	019.00	0
019	Calcium (%)	2196	1.190	0.0000	1.184	0.0580	0.0244	122	0.11	019.31	0
019	Calcium (%)	0100	1.190	0.0600	1.184	0.0580	0.0244	122	0.11	019.99	0
019	Calcium (%)	0407	1.190	0.0149	1.184	0.0580	0.0244	122	0.11	019.41	0
019	Calcium (%)	0504	1.191	0.0300	1.184	0.0580	0.0244	122	0.12	019.42	0
019	Calcium (%)	2144	1.195	0.0100	1.184	0.0580	0.0244	122	0.19	019.00	0
019	Calcium (%)	0098	1.195	0.0100	1.184	0.0580	0.0244	122	0.19	019.41	0
019	Calcium (%)	2144	1.195	0.0100	1.184	0.0580	0.0244	122	0.19	019.99	0
019	Calcium (%)	2200	1.200	0.0200	1.184	0.0580	0.0244	122	0.28	019.08	0
019	Calcium (%)	2012	1.200	0.0000	1.184	0.0580	0.0244	122	0.28	019.41	0
019	Calcium (%)	0598	1.200	0.0158	1.184	0.0580	0.0244	122	0.28	019.41	0
019	Calcium (%)	2314	1.205	0.0002	1.184	0.0580	0.0244	122	0.36	019.43	0
019	Calcium (%)	2302	1.205	0.0100	1.184	0.0580	0.0244	122	0.37	019.00	0
019	Calcium (%)	0687	1.205	0.0500	1.184	0.0580	0.0244	122	0.37	019.31	0
019	Calcium (%)	0049	1.205	0.0300	1.184	0.0580	0.0244	122	0.37	019.41	0
019	Calcium (%)	0148	1.205	0.0500	1.184	0.0580	0.0244	122	0.37	019.41	0
019	Calcium (%)	2089	1.205	0.0100	1.184	0.0580	0.0244	122	0.37	019.43	0
019	Calcium (%)	2022	1.210	0.0200	1.184	0.0580	0.0244	122	0.45	019.00	0
019	Calcium (%)	0017	1.213	0.0091	1.184	0.0580	0.0244	122	0.50	019.43	0
019	Calcium (%)	2337	1.215	0.0100	1.184	0.0580	0.0244	122	0.54	019.08	0
019	Calcium (%)	0723	1.216	0.0020	1.184	0.0580	0.0244	122	0.55	019.41	0
019	Calcium (%)	0683	1.220	0.0400	1.184	0.0580	0.0244	122	0.62	019.00	0
019	Calcium (%)	0505	1.220	0.0800	1.184	0.0580	0.0244	122	0.62	019.43	0
019	Calcium (%)	0939	1.225	0.0500	1.184	0.0580	0.0244	122	0.71	019.31	0
019	Calcium (%)	0042	1.225	0.0300	1.184	0.0580	0.0244	122	0.71	019.42	0
019	Calcium (%)	0265	1.225	0.0300	1.184	0.0580	0.0244	122	0.71	019.42	0
019	Calcium (%)	0610	1.225	0.0700	1.184	0.0580	0.0244	122	0.71	019.43	0
019	Calcium (%)	0142	1.225	0.0106	1.184	0.0580	0.0244	122	0.72	019.31	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
019	Calcium (%)	0689	1.230	0.0200	1.184	0.0580	0.0244	122	0.80	019.08	0
019	Calcium (%)	0682	1.230	0.0000	1.184	0.0580	0.0244	122	0.80	019.43	0
019	Calcium (%)	0553	1.230	0.0000	1.184	0.0580	0.0244	122	0.80	019.53	0
019	Calcium (%)	2259	1.232	0.0400	1.184	0.0580	0.0244	122	0.83	019.41	0
019	Calcium (%)	0171	1.235	0.0100	1.184	0.0580	0.0244	122	0.88	019.41	0
019	Calcium (%)	0366	1.235	0.0500	1.184	0.0580	0.0244	122	0.88	019.42	0
019	Calcium (%)	0354	1.236	0.0030	1.184	0.0580	0.0244	122	0.89	019.41	0
019	Calcium (%)	2394	1.237	0.0000	1.184	0.0580	0.0244	122	0.92	019.43	0
019	Calcium (%)	0186	1.240	0.0068	1.184	0.0580	0.0244	122	0.96	019.42	0
019	Calcium (%)	0510	1.240	0.0400	1.184	0.0580	0.0244	122	0.97	019.43	0
019	Calcium (%)	2190	1.241	0.0205	1.184	0.0580	0.0244	122	0.99	019.08	0
019	Calcium (%)	0300	1.250	0.1000	1.184	0.0580	0.0244	122	1.14	019.43	0
019	Calcium (%)	0353	1.250	0.0200	1.184	0.0580	0.0244	122	1.14	019.43	0
019	Calcium (%)	2022	1.260	0.0600	1.184	0.0580	0.0244	122	1.31	019.33	0
019	Calcium (%)	0626	1.267	0.0170	1.184	0.0580	0.0244	122	1.43	019.43	0
019	Calcium (%)	0027	1.273	0.0130	1.184	0.0580	0.0244	122	1.53	019.43	0
019	Calcium (%)	0885	1.274	0.0070	1.184	0.0580	0.0244	122	1.55	019.08	0
019	Calcium (%)	0186	1.279	0.0154	1.184	0.0580	0.0244	122	1.64	019.52	0
019	Calcium (%)	0098	1.280	0.0000	1.184	0.0580	0.0244	122	1.66	019.43	0
019	Calcium (%)	0108	1.295	0.0300	1.184	0.0580	0.0244	122	1.92	019.31	0
019	Calcium (%)	0536	1.335	0.0100	1.184	0.0580	0.0244	122	2.61	019.02	0
019	Calcium (%)	0278	1.345	0.0500	1.184	0.0580	0.0244	122	2.78	019.42	0
019	Calcium (%)	2397	1.350	0.0200	1.184	0.0580	0.0244	122	2.86	019.08	0
019	Calcium (%)	0190	1.370	0.0000	1.184	0.0580	0.0244	122	3.21	019.42	0
019	Calcium (%)	0009	1.387	0.0072	1.184	0.0580	0.0244	122	3.50	019.42	0
019	Calcium (%)	0726	1.445	0.0300	1.184	0.0580	0.0244	122	4.50	019.42	0
019	Calcium (%)	2326	1.450	0.1000	1.184	0.0580	0.0244	122	4.59	019.43	0
019	Calcium (%)	0045	1.067	0.2670	1.184	0.0580	0.0244	122	-2.02	019.42	1
019	Calcium (%)	0692	1.255	0.1500	1.184	0.0580	0.0244	122	1.23	019.42	1
019	Calcium (%)	2375	1.505	0.1500	1.184	0.0580	0.0244	122	5.54	019.42	1
021	Cobalt (ppm)	0619	0.7910	0.0020	1.465	0.3144	0.0760	20	-2.14	021.41	0
021	Cobalt (ppm)	0345	1.020	0.0600	1.465	0.3144	0.0760	20	-1.41	021.43	0
021	Cobalt (ppm)	0123	1.090	0.0000	1.465	0.3144	0.0760	20	-1.19	021.99	0
021	Cobalt (ppm)	0164	1.200	0.2000	1.465	0.3144	0.0760	20	-0.84	021.41	0
021	Cobalt (ppm)	0171	1.265	0.0300	1.465	0.3144	0.0760	20	-0.63	021.41	0
021	Cobalt (ppm)	0045	1.325	0.0100	1.465	0.3144	0.0760	20	-0.44	021.42	0
021	Cobalt (ppm)	2326	1.350	0.2400	1.465	0.3144	0.0760	20	-0.36	021.43	0
021	Cobalt (ppm)	0682	1.400	0.0000	1.465	0.3144	0.0760	20	-0.21	021.43	0
021	Cobalt (ppm)	0366	1.425	0.0500	1.465	0.3144	0.0760	20	-0.13	021.42	0
021	Cobalt (ppm)	0610	1.430	0.0400	1.465	0.3144	0.0760	20	-0.11	021.53	0
021	Cobalt (ppm)	0407	1.433	0.1003	1.465	0.3144	0.0760	20	-0.10	021.41	0
021	Cobalt (ppm)	0629	1.480	0.0200	1.465	0.3144	0.0760	20	0.05	021.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
021	Cobalt (ppm)	0407	1.550	0.0649	1.465	0.3144	0.0760	20	0.27	021.53	0
021	Cobalt (ppm)	0553	1.560	0.1800	1.465	0.3144	0.0760	20	0.30	021.53	0
021	Cobalt (ppm)	0510	1.675	0.0100	1.465	0.3144	0.0760	20	0.67	021.43	0
021	Cobalt (ppm)	0407	1.689	0.0080	1.465	0.3144	0.0760	20	0.71	021.43	0
021	Cobalt (ppm)	0148	1.695	0.1300	1.465	0.3144	0.0760	20	0.73	021.41	0
021	Cobalt (ppm)	0169	1.840	0.1200	1.465	0.3144	0.0760	20	1.19	021.43	0
021	Cobalt (ppm)	0186	1.945	0.0700	1.465	0.3144	0.0760	20	1.53	021.52	0
021	Cobalt (ppm)	0066	1.987	0.1840	1.465	0.3144	0.0760	20	1.66	021.41	0
022	Copper (ppm)	0366	0.2450	0.0300	23.34	2.504	0.9015	84	-9.23	022.42	0
022	Copper (ppm)	2022	14.56	1.740	23.34	2.504	0.9015	84	-3.51	022.33	0
022	Copper (ppm)	0656	15.77	0.0000	23.34	2.504	0.9015	84	-3.03	022.31	0
022	Copper (ppm)	0242	17.50	1.000	23.34	2.504	0.9015	84	-2.33	022.99	0
022	Copper (ppm)	2326	17.82	1.130	23.34	2.504	0.9015	84	-2.21	022.43	0
022	Copper (ppm)	0598	18.14	1.668	23.34	2.504	0.9015	84	-2.08	022.41	0
022	Copper (ppm)	0921	18.21	4.720	23.34	2.504	0.9015	84	-2.05	022.31	0
022	Copper (ppm)	0294	18.28	0.4400	23.34	2.504	0.9015	84	-2.02	022.42	0
022	Copper (ppm)	0629	18.50	1.000	23.34	2.504	0.9015	84	-1.93	022.31	0
022	Copper (ppm)	0848	19.62	0.0500	23.34	2.504	0.9015	84	-1.49	022.41	0
022	Copper (ppm)	0674	20.08	0.0000	23.34	2.504	0.9015	84	-1.30	022.41	0
022	Copper (ppm)	0229	20.51	0.4200	23.34	2.504	0.9015	84	-1.13	022.41	0
022	Copper (ppm)	0049	20.59	1.640	23.34	2.504	0.9015	84	-1.10	022.41	0
022	Copper (ppm)	0001	20.97	0.0900	23.34	2.504	0.9015	84	-0.95	022.31	0
022	Copper (ppm)	0226	21.22	0.4500	23.34	2.504	0.9015	84	-0.85	022.43	0
022	Copper (ppm)	0358	21.22	0.6000	23.34	2.504	0.9015	84	-0.85	022.41	0
022	Copper (ppm)	0619	21.25	1.300	23.34	2.504	0.9015	84	-0.84	022.41	0
022	Copper (ppm)	0354	21.31	0.0300	23.34	2.504	0.9015	84	-0.81	022.41	0
022	Copper (ppm)	0353	21.50	1.000	23.34	2.504	0.9015	84	-0.74	022.43	0
022	Copper (ppm)	0723	21.66	1.388	23.34	2.504	0.9015	84	-0.67	022.41	0
022	Copper (ppm)	0520	21.87	0.6320	23.34	2.504	0.9015	84	-0.59	022.41	0
022	Copper (ppm)	0407	21.94	0.4363	23.34	2.504	0.9015	84	-0.56	022.43	0
022	Copper (ppm)	0511	22.00	0.0000	23.34	2.504	0.9015	84	-0.54	022.43	0
022	Copper (ppm)	0100	22.00	2.000	23.34	2.504	0.9015	84	-0.54	022.99	0
022	Copper (ppm)	0870	22.12	0.8000	23.34	2.504	0.9015	84	-0.49	022.42	0
022	Copper (ppm)	0202	22.20	3.000	23.34	2.504	0.9015	84	-0.46	022.43	0
022	Copper (ppm)	0004	22.26	0.1800	23.34	2.504	0.9015	84	-0.43	022.41	0
022	Copper (ppm)	0870	22.42	0.2300	23.34	2.504	0.9015	84	-0.37	022.43	0
022	Copper (ppm)	0066	22.46	0.7900	23.34	2.504	0.9015	84	-0.36	022.44	0
022	Copper (ppm)	0297	22.50	3.000	23.34	2.504	0.9015	84	-0.34	022.43	0
022	Copper (ppm)	0164	22.50	1.000	23.34	2.504	0.9015	84	-0.34	022.44	0
022	Copper (ppm)	0553	22.55	1.700	23.34	2.504	0.9015	84	-0.32	022.53	0
022	Copper (ppm)	0910	23.00	0.0000	23.34	2.504	0.9015	84	-0.14	022.41	0
022	Copper (ppm)	0148	23.14	0.1424	23.34	2.504	0.9015	84	-0.08	022.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
022	Copper (ppm)	0098	23.16	0.4500	23.34	2.504	0.9015	84	-0.08	022.41	0
022	Copper (ppm)	0675	23.22	0.0300	23.34	2.504	0.9015	84	-0.05	022.43	0
022	Copper (ppm)	0169	23.35	0.3000	23.34	2.504	0.9015	84	0.00	022.43	0
022	Copper (ppm)	0629	23.35	0.5000	23.34	2.504	0.9015	84	0.00	022.43	0
022	Copper (ppm)	0848	23.44	0.8500	23.34	2.504	0.9015	84	0.04	022.43	0
022	Copper (ppm)	0407	23.48	2.127	23.34	2.504	0.9015	84	0.05	022.53	0
022	Copper (ppm)	0033	23.55	0.3000	23.34	2.504	0.9015	84	0.08	022.43	0
022	Copper (ppm)	2129	23.57	1.090	23.34	2.504	0.9015	84	0.09	022.42	0
022	Copper (ppm)	0968	23.59	0.0390	23.34	2.504	0.9015	84	0.10	022.43	0
022	Copper (ppm)	0019	23.75	0.4200	23.34	2.504	0.9015	84	0.16	022.41	0
022	Copper (ppm)	0345	23.76	0.0800	23.34	2.504	0.9015	84	0.17	022.43	0
022	Copper (ppm)	2089	23.79	2.140	23.34	2.504	0.9015	84	0.18	022.43	0
022	Copper (ppm)	0263	23.82	0.1200	23.34	2.504	0.9015	84	0.19	022.42	0
022	Copper (ppm)	0083	23.89	0.3300	23.34	2.504	0.9015	84	0.22	022.43	0
022	Copper (ppm)	0027	23.92	1.977	23.34	2.504	0.9015	84	0.23	022.43	0
022	Copper (ppm)	2196	23.92	0.0000	23.34	2.504	0.9015	84	0.23	022.31	0
022	Copper (ppm)	0939	23.98	0.5600	23.34	2.504	0.9015	84	0.25	022.31	0
022	Copper (ppm)	0626	23.99	0.3700	23.34	2.504	0.9015	84	0.26	022.43	0
022	Copper (ppm)	0010	24.00	1.400	23.34	2.504	0.9015	84	0.26	022.33	0
022	Copper (ppm)	0035	24.00	0.0000	23.34	2.504	0.9015	84	0.26	022.42	0
022	Copper (ppm)	0510	24.00	0.0000	23.34	2.504	0.9015	84	0.26	022.43	0
022	Copper (ppm)	2319	24.10	0.6000	23.34	2.504	0.9015	84	0.30	022.42	0
022	Copper (ppm)	0017	24.11	0.4524	23.34	2.504	0.9015	84	0.30	022.43	0
022	Copper (ppm)	0948	24.13	3.280	23.34	2.504	0.9015	84	0.31	022.33	0
022	Copper (ppm)	0407	24.14	0.3918	23.34	2.504	0.9015	84	0.32	022.41	0
022	Copper (ppm)	0726	24.25	0.1000	23.34	2.504	0.9015	84	0.36	022.42	0
022	Copper (ppm)	0357	24.40	0.0560	23.34	2.504	0.9015	84	0.42	022.42	0
022	Copper (ppm)	0098	24.42	1.360	23.34	2.504	0.9015	84	0.43	022.43	0
022	Copper (ppm)	0226	24.43	0.7200	23.34	2.504	0.9015	84	0.43	022.41	0
022	Copper (ppm)	0505	25.15	3.010	23.34	2.504	0.9015	84	0.72	022.43	0
022	Copper (ppm)	2012	25.24	0.7200	23.34	2.504	0.9015	84	0.76	022.41	0
022	Copper (ppm)	0045	25.35	0.1000	23.34	2.504	0.9015	84	0.80	022.42	0
022	Copper (ppm)	0047	25.52	0.6800	23.34	2.504	0.9015	84	0.87	022.52	0
022	Copper (ppm)	0171	25.55	0.0200	23.34	2.504	0.9015	84	0.88	022.41	0
022	Copper (ppm)	0610	25.70	0.4000	23.34	2.504	0.9015	84	0.94	022.43	0
022	Copper (ppm)	0693	25.73	3.440	23.34	2.504	0.9015	84	0.95	022.42	0
022	Copper (ppm)	0123	25.80	0.0000	23.34	2.504	0.9015	84	0.98	022.99	0
022	Copper (ppm)	0035	26.00	0.0000	23.34	2.504	0.9015	84	1.06	022.43	0
022	Copper (ppm)	2192	26.00	0.0000	23.34	2.504	0.9015	84	1.06	022.43	0
022	Copper (ppm)	0190	26.04	0.0100	23.34	2.504	0.9015	84	1.07	022.42	0
022	Copper (ppm)	2188	26.08	0.0000	23.34	2.504	0.9015	84	1.09	022.31	0
022	Copper (ppm)	0265	26.10	3.400	23.34	2.504	0.9015	84	1.10	022.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
022	Copper (ppm)	0682	26.70	0.0000	23.34	2.504	0.9015	84	1.34	022.43	0
022	Copper (ppm)	0504	26.75	1.100	23.34	2.504	0.9015	84	1.36	022.42	0
022	Copper (ppm)	0651	27.69	0.9977	23.34	2.504	0.9015	84	1.73	022.31	0
022	Copper (ppm)	0042	27.75	1.500	23.34	2.504	0.9015	84	1.76	022.42	0
022	Copper (ppm)	0009	28.15	3.300	23.34	2.504	0.9015	84	1.92	022.42	0
022	Copper (ppm)	2375	29.00	0.0000	23.34	2.504	0.9015	84	2.26	022.42	0
022	Copper (ppm)	0529	31.50	1.000	23.34	2.504	0.9015	84	3.26	022.31	0
022	Copper (ppm)	0278	31.80	3.400	23.34	2.504	0.9015	84	3.38	022.42	0
022	Copper (ppm)	0692	22.45	6.500	23.34	2.504	0.9015	84	-0.36	022.42	1
022	Copper (ppm)	2246	226.8	6.730	23.34	2.504	0.9015	84	81.26	022.31	2
025	Iron (ppm)	0723	182.3	3.153	227.1	18.38	5.830	80	-2.44	025.41	0
025	Iron (ppm)	0265	182.5	15.00	227.1	18.38	5.830	80	-2.42	025.42	0
025	Iron (ppm)	2326	184.8	9.200	227.1	18.38	5.830	80	-2.30	025.43	0
025	Iron (ppm)	0619	185.0	14.00	227.1	18.38	5.830	80	-2.29	025.41	0
025	Iron (ppm)	0066	194.1	5.571	227.1	18.38	5.830	80	-1.80	025.41	0
025	Iron (ppm)	2319	195.5	5.000	227.1	18.38	5.830	80	-1.72	025.42	0
025	Iron (ppm)	0226	198.4	6.390	227.1	18.38	5.830	80	-1.56	025.43	0
025	Iron (ppm)	0294	199.5	0.5000	227.1	18.38	5.830	80	-1.50	025.42	0
025	Iron (ppm)	0629	200.0	4.000	227.1	18.38	5.830	80	-1.47	025.31	0
025	Iron (ppm)	0226	201.8	14.15	227.1	18.38	5.830	80	-1.37	025.41	0
025	Iron (ppm)	0242	202.5	1.000	227.1	18.38	5.830	80	-1.34	025.99	0
025	Iron (ppm)	2089	207.9	4.890	227.1	18.38	5.830	80	-1.04	025.43	0
025	Iron (ppm)	0100	209.0	2.000	227.1	18.38	5.830	80	-0.98	025.99	0
025	Iron (ppm)	0345	209.2	1.900	227.1	18.38	5.830	80	-0.98	025.43	0
025	Iron (ppm)	0123	210.0	0.0000	227.1	18.38	5.830	80	-0.93	025.99	0
025	Iron (ppm)	2246	210.6	1.520	227.1	18.38	5.830	80	-0.90	025.31	0
025	Iron (ppm)	0520	211.8	5.920	227.1	18.38	5.830	80	-0.83	025.41	0
025	Iron (ppm)	0553	212.5	1.000	227.1	18.38	5.830	80	-0.79	025.53	0
025	Iron (ppm)	0529	214.5	11.00	227.1	18.38	5.830	80	-0.68	025.31	0
025	Iron (ppm)	0083	214.9	1.470	227.1	18.38	5.830	80	-0.66	025.43	0
025	Iron (ppm)	0511	216.5	7.000	227.1	18.38	5.830	80	-0.58	025.43	0
025	Iron (ppm)	0354	217.1	0.0600	227.1	18.38	5.830	80	-0.54	025.41	0
025	Iron (ppm)	0297	218.0	26.00	227.1	18.38	5.830	80	-0.49	025.43	0
025	Iron (ppm)	0358	219.2	8.050	227.1	18.38	5.830	80	-0.43	025.41	0
025	Iron (ppm)	0190	219.6	1.100	227.1	18.38	5.830	80	-0.41	025.42	0
025	Iron (ppm)	0598	221.2	1.444	227.1	18.38	5.830	80	-0.32	025.41	0
025	Iron (ppm)	0726	222.8	10.20	227.1	18.38	5.830	80	-0.23	025.42	0
025	Iron (ppm)	0047	222.9	1.000	227.1	18.38	5.830	80	-0.23	025.52	0
025	Iron (ppm)	0019	223.3	4.400	227.1	18.38	5.830	80	-0.21	025.41	0
025	Iron (ppm)	0042	223.5	3.000	227.1	18.38	5.830	80	-0.19	025.42	0
025	Iron (ppm)	0164	224.5	1.000	227.1	18.38	5.830	80	-0.14	025.41	0
025	Iron (ppm)	0169	224.5	3.000	227.1	18.38	5.830	80	-0.14	025.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
025	Iron (ppm)	0407	225.8	8.350	227.1	18.38	5.830	80	-0.07	025.41	0
025	Iron (ppm)	0407	226.8	2.938	227.1	18.38	5.830	80	-0.02	025.43	0
025	Iron (ppm)	0045	227.0	2.000	227.1	18.38	5.830	80	0.00	025.42	0
025	Iron (ppm)	0968	227.1	3.816	227.1	18.38	5.830	80	0.00	025.43	0
025	Iron (ppm)	0675	227.5	1.870	227.1	18.38	5.830	80	0.02	025.43	0
025	Iron (ppm)	2196	227.6	0.0000	227.1	18.38	5.830	80	0.03	025.31	0
025	Iron (ppm)	0610	227.8	13.50	227.1	18.38	5.830	80	0.04	025.43	0
025	Iron (ppm)	0505	227.9	4.900	227.1	18.38	5.830	80	0.04	025.43	0
025	Iron (ppm)	0629	228.0	0.0000	227.1	18.38	5.830	80	0.05	025.43	0
025	Iron (ppm)	0229	228.1	2.300	227.1	18.38	5.830	80	0.05	025.41	0
025	Iron (ppm)	2012	228.8	6.500	227.1	18.38	5.830	80	0.09	025.41	0
025	Iron (ppm)	0910	229.0	8.000	227.1	18.38	5.830	80	0.10	025.41	0
025	Iron (ppm)	0939	229.6	2.460	227.1	18.38	5.830	80	0.14	025.31	0
025	Iron (ppm)	0098	230.0	5.200	227.1	18.38	5.830	80	0.16	025.41	0
025	Iron (ppm)	0407	230.1	0.6366	227.1	18.38	5.830	80	0.17	025.53	0
025	Iron (ppm)	0357	230.7	1.425	227.1	18.38	5.830	80	0.20	025.42	0
025	Iron (ppm)	0512	231.2	1.800	227.1	18.38	5.830	80	0.22	025.43	0
025	Iron (ppm)	0263	231.8	21.32	227.1	18.38	5.830	80	0.26	025.42	0
025	Iron (ppm)	0656	232.2	0.0000	227.1	18.38	5.830	80	0.28	025.31	0
025	Iron (ppm)	0693	232.2	8.270	227.1	18.38	5.830	80	0.28	025.42	0
025	Iron (ppm)	0033	233.0	16.00	227.1	18.38	5.830	80	0.32	025.43	0
025	Iron (ppm)	0626	233.3	22.55	227.1	18.38	5.830	80	0.34	025.43	0
025	Iron (ppm)	0674	233.3	0.0000	227.1	18.38	5.830	80	0.34	025.41	0
025	Iron (ppm)	2192	234.0	8.000	227.1	18.38	5.830	80	0.38	025.43	0
025	Iron (ppm)	0682	235.1	0.0000	227.1	18.38	5.830	80	0.43	025.43	0
025	Iron (ppm)	0049	235.5	9.020	227.1	18.38	5.830	80	0.46	025.41	0
025	Iron (ppm)	0035	235.5	1.000	227.1	18.38	5.830	80	0.46	025.42	0
025	Iron (ppm)	0171	238.2	5.350	227.1	18.38	5.830	80	0.60	025.41	0
025	Iron (ppm)	0004	238.8	4.970	227.1	18.38	5.830	80	0.64	025.41	0
025	Iron (ppm)	0504	240.0	6.000	227.1	18.38	5.830	80	0.70	025.42	0
025	Iron (ppm)	2375	240.5	29.00	227.1	18.38	5.830	80	0.73	025.42	0
025	Iron (ppm)	0948	240.9	5.880	227.1	18.38	5.830	80	0.75	025.33	0
025	Iron (ppm)	0510	241.5	3.000	227.1	18.38	5.830	80	0.78	025.43	0
025	Iron (ppm)	0001	241.9	1.600	227.1	18.38	5.830	80	0.81	025.31	0
025	Iron (ppm)	0870	245.2	0.4000	227.1	18.38	5.830	80	0.99	025.42	0
025	Iron (ppm)	0098	245.7	13.70	227.1	18.38	5.830	80	1.01	025.43	0
025	Iron (ppm)	0870	246.1	11.00	227.1	18.38	5.830	80	1.03	025.43	0
025	Iron (ppm)	0017	246.3	4.731	227.1	18.38	5.830	80	1.05	025.43	0
025	Iron (ppm)	2022	246.7	1.000	227.1	18.38	5.830	80	1.07	025.33	0
025	Iron (ppm)	0353	250.0	2.000	227.1	18.38	5.830	80	1.25	025.43	0
025	Iron (ppm)	0035	250.5	3.000	227.1	18.38	5.830	80	1.27	025.43	0
025	Iron (ppm)	0876	254.0	28.00	227.1	18.38	5.830	80	1.46	025.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
025	Iron (ppm)	0148	254.0	5.805	227.1	18.38	5.830	80	1.46	025.41	0
025	Iron (ppm)	2129	261.4	5.000	227.1	18.38	5.830	80	1.87	025.42	0
025	Iron (ppm)	0651	262.7	1.164	227.1	18.38	5.830	80	1.94	025.31	0
025	Iron (ppm)	2188	264.0	0.0000	227.1	18.38	5.830	80	2.01	025.31	0
025	Iron (ppm)	0366	265.0	2.000	227.1	18.38	5.830	80	2.06	025.42	0
025	Iron (ppm)	0038	292.5	7.000	227.1	18.38	5.830	80	3.56	025.31	0
025	Iron (ppm)	0692	208.5	51.00	227.1	18.38	5.830	80	-1.01	025.42	1
025	Iron (ppm)	0278	229.5	39.00	227.1	18.38	5.830	80	0.13	025.42	1
027	Magnesium (%)	0169	0.2300	0.0200	0.3061	0.0144	0.0064	87	-5.28	027.32	0
027	Magnesium (%)	2326	0.2350	0.0100	0.3061	0.0144	0.0064	87	-4.93	027.43	0
027	Magnesium (%)	0226	0.2700	0.0000	0.3061	0.0144	0.0064	87	-2.50	027.43	0
027	Magnesium (%)	0407	0.2800	0.0032	0.3061	0.0144	0.0064	87	-1.81	027.43	0
027	Magnesium (%)	2192	0.2850	0.0100	0.3061	0.0144	0.0064	87	-1.46	027.43	0
027	Magnesium (%)	0066	0.2860	0.0220	0.3061	0.0144	0.0064	87	-1.39	027.41	0
027	Magnesium (%)	2129	0.2874	0.0021	0.3061	0.0144	0.0064	87	-1.30	027.42	0
027	Magnesium (%)	0083	0.2895	0.0010	0.3061	0.0144	0.0064	87	-1.15	027.43	0
027	Magnesium (%)	0629	0.2900	0.0000	0.3061	0.0144	0.0064	87	-1.12	027.31	0
027	Magnesium (%)	2246	0.2900	0.0000	0.3061	0.0144	0.0064	87	-1.12	027.31	0
027	Magnesium (%)	0004	0.2900	0.0000	0.3061	0.0144	0.0064	87	-1.12	027.41	0
027	Magnesium (%)	0229	0.2900	0.0000	0.3061	0.0144	0.0064	87	-1.12	027.41	0
027	Magnesium (%)	0910	0.2900	0.0000	0.3061	0.0144	0.0064	87	-1.12	027.41	0
027	Magnesium (%)	0089	0.2900	0.0000	0.3061	0.0144	0.0064	87	-1.12	027.43	0
027	Magnesium (%)	0629	0.2900	0.0000	0.3061	0.0144	0.0064	87	-1.12	027.43	0
027	Magnesium (%)	0263	0.2925	0.0050	0.3061	0.0144	0.0064	87	-0.94	027.42	0
027	Magnesium (%)	0006	0.2930	0.0020	0.3061	0.0144	0.0064	87	-0.91	027.43	0
027	Magnesium (%)	0512	0.2947	0.0039	0.3061	0.0144	0.0064	87	-0.79	027.43	0
027	Magnesium (%)	0265	0.2950	0.0100	0.3061	0.0144	0.0064	87	-0.77	027.42	0
027	Magnesium (%)	0510	0.2950	0.0100	0.3061	0.0144	0.0064	87	-0.77	027.43	0
027	Magnesium (%)	0123	0.2950	0.0100	0.3061	0.0144	0.0064	87	-0.77	027.99	0
027	Magnesium (%)	0407	0.2960	0.0083	0.3061	0.0144	0.0064	87	-0.70	027.41	0
027	Magnesium (%)	0142	0.2978	0.0074	0.3061	0.0144	0.0064	87	-0.58	027.31	0
027	Magnesium (%)	0968	0.2980	0.0020	0.3061	0.0144	0.0064	87	-0.56	027.43	0
027	Magnesium (%)	0345	0.2985	0.0030	0.3061	0.0144	0.0064	87	-0.53	027.43	0
027	Magnesium (%)	0870	0.2987	0.0116	0.3061	0.0144	0.0064	87	-0.51	027.42	0
027	Magnesium (%)	0001	0.2989	0.0045	0.3061	0.0144	0.0064	87	-0.50	027.31	0
027	Magnesium (%)	0019	0.3000	0.0000	0.3061	0.0144	0.0064	87	-0.42	027.41	0
027	Magnesium (%)	0226	0.3000	0.0000	0.3061	0.0144	0.0064	87	-0.42	027.41	0
027	Magnesium (%)	0358	0.3000	0.0000	0.3061	0.0144	0.0064	87	-0.42	027.41	0
027	Magnesium (%)	2319	0.3000	0.0000	0.3061	0.0144	0.0064	87	-0.42	027.42	0
027	Magnesium (%)	0033	0.3000	0.0200	0.3061	0.0144	0.0064	87	-0.42	027.43	0
027	Magnesium (%)	0511	0.3000	0.0000	0.3061	0.0144	0.0064	87	-0.42	027.43	0
027	Magnesium (%)	0610	0.3000	0.0000	0.3061	0.0144	0.0064	87	-0.42	027.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
027	Magnesium (%)	0242	0.3000	0.0000	0.3061	0.0144	0.0064	87	-0.42	027.99	0
027	Magnesium (%)	0407	0.3035	0.0065	0.3061	0.0144	0.0064	87	-0.18	027.53	0
027	Magnesium (%)	0357	0.3037	0.0006	0.3061	0.0144	0.0064	87	-0.17	027.42	0
027	Magnesium (%)	0164	0.3040	0.0000	0.3061	0.0144	0.0064	87	-0.15	027.44	0
027	Magnesium (%)	0049	0.3050	0.0300	0.3061	0.0144	0.0064	87	-0.08	027.41	0
027	Magnesium (%)	0294	0.3050	0.0100	0.3061	0.0144	0.0064	87	-0.08	027.42	0
027	Magnesium (%)	0297	0.3050	0.0100	0.3061	0.0144	0.0064	87	-0.08	027.43	0
027	Magnesium (%)	0505	0.3050	0.0100	0.3061	0.0144	0.0064	87	-0.08	027.43	0
027	Magnesium (%)	0675	0.3050	0.0100	0.3061	0.0144	0.0064	87	-0.08	027.43	0
027	Magnesium (%)	0148	0.3055	0.0110	0.3061	0.0144	0.0064	87	-0.04	027.41	0
027	Magnesium (%)	0202	0.3055	0.0130	0.3061	0.0144	0.0064	87	-0.04	027.43	0
027	Magnesium (%)	0504	0.3060	0.0040	0.3061	0.0144	0.0064	87	-0.01	027.42	0
027	Magnesium (%)	0948	0.3080	0.0020	0.3061	0.0144	0.0064	87	0.13	027.33	0
027	Magnesium (%)	0042	0.3080	0.0040	0.3061	0.0144	0.0064	87	0.13	027.42	0
027	Magnesium (%)	0619	0.3085	0.0070	0.3061	0.0144	0.0064	87	0.17	027.41	0
027	Magnesium (%)	2012	0.3085	0.0070	0.3061	0.0144	0.0064	87	0.17	027.41	0
027	Magnesium (%)	0870	0.3086	0.0185	0.3061	0.0144	0.0064	87	0.17	027.43	0
027	Magnesium (%)	0529	0.3100	0.0000	0.3061	0.0144	0.0064	87	0.27	027.31	0
027	Magnesium (%)	0656	0.3100	0.0000	0.3061	0.0144	0.0064	87	0.27	027.31	0
027	Magnesium (%)	0098	0.3100	0.0000	0.3061	0.0144	0.0064	87	0.27	027.41	0
027	Magnesium (%)	0171	0.3100	0.0000	0.3061	0.0144	0.0064	87	0.27	027.41	0
027	Magnesium (%)	0190	0.3100	0.0000	0.3061	0.0144	0.0064	87	0.27	027.42	0
027	Magnesium (%)	0726	0.3100	0.0000	0.3061	0.0144	0.0064	87	0.27	027.42	0
027	Magnesium (%)	0098	0.3100	0.0200	0.3061	0.0144	0.0064	87	0.27	027.43	0
027	Magnesium (%)	2089	0.3100	0.0000	0.3061	0.0144	0.0064	87	0.27	027.43	0
027	Magnesium (%)	0553	0.3105	0.0030	0.3061	0.0144	0.0064	87	0.30	027.53	0
027	Magnesium (%)	0047	0.3110	0.0140	0.3061	0.0144	0.0064	87	0.34	027.52	0
027	Magnesium (%)	0038	0.3135	0.0070	0.3061	0.0144	0.0064	87	0.51	027.31	0
027	Magnesium (%)	0027	0.3135	0.0030	0.3061	0.0144	0.0064	87	0.51	027.43	0
027	Magnesium (%)	0939	0.3150	0.0100	0.3061	0.0144	0.0064	87	0.62	027.31	0
027	Magnesium (%)	0100	0.3150	0.0100	0.3061	0.0144	0.0064	87	0.62	027.99	0
027	Magnesium (%)	0598	0.3153	0.0020	0.3061	0.0144	0.0064	87	0.64	027.41	0
027	Magnesium (%)	2146	0.3200	0.0000	0.3061	0.0144	0.0064	87	0.96	027.31	0
027	Magnesium (%)	2196	0.3200	0.0000	0.3061	0.0144	0.0064	87	0.96	027.31	0
027	Magnesium (%)	0035	0.3200	0.0000	0.3061	0.0144	0.0064	87	0.96	027.42	0
027	Magnesium (%)	0366	0.3200	0.0200	0.3061	0.0144	0.0064	87	0.96	027.42	0
027	Magnesium (%)	0035	0.3200	0.0000	0.3061	0.0144	0.0064	87	0.96	027.43	0
027	Magnesium (%)	0682	0.3200	0.0000	0.3061	0.0144	0.0064	87	0.96	027.43	0
027	Magnesium (%)	2146	0.3200	0.0000	0.3061	0.0144	0.0064	87	0.96	027.51	0
027	Magnesium (%)	2144	0.3200	0.0000	0.3061	0.0144	0.0064	87	0.96	027.99	0
027	Magnesium (%)	0186	0.3213	0.0003	0.3061	0.0144	0.0064	87	1.05	027.52	0
027	Magnesium (%)	0876	0.3215	0.0130	0.3061	0.0144	0.0064	87	1.07	027.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
027	Magnesium (%)	0354	0.3230	0.0000	0.3061	0.0144	0.0064	87	1.17	027.41	0
027	Magnesium (%)	0186	0.3231	0.0003	0.3061	0.0144	0.0064	87	1.18	027.42	0
027	Magnesium (%)	0693	0.3235	0.0370	0.3061	0.0144	0.0064	87	1.21	027.42	0
027	Magnesium (%)	0520	0.3275	0.0130	0.3061	0.0144	0.0064	87	1.48	027.41	0
027	Magnesium (%)	0017	0.3276	0.0002	0.3061	0.0144	0.0064	87	1.49	027.43	0
027	Magnesium (%)	0674	0.3300	0.0000	0.3061	0.0144	0.0064	87	1.66	027.41	0
027	Magnesium (%)	0353	0.3350	0.0100	0.3061	0.0144	0.0064	87	2.00	027.43	0
027	Magnesium (%)	0009	0.3438	0.0270	0.3061	0.0144	0.0064	87	2.61	027.42	0
027	Magnesium (%)	0626	0.3510	0.0080	0.3061	0.0144	0.0064	87	3.11	027.43	0
027	Magnesium (%)	0278	0.3700	0.0200	0.3061	0.0144	0.0064	87	4.43	027.42	0
027	Magnesium (%)	2375	0.3750	0.0300	0.3061	0.0144	0.0064	87	4.78	027.42	0
027	Magnesium (%)	0045	0.2790	0.0520	0.3061	0.0144	0.0064	87	-1.88	027.42	1
027	Magnesium (%)	0692	0.3000	0.0600	0.3061	0.0144	0.0064	87	-0.42	027.42	1
028	Manganese (ppm)	0066	95.99	2.600	128.3	7.933	4.158	85	-4.07	028.44	0
028	Manganese (ppm)	0723	102.9	5.323	128.3	7.933	4.158	85	-3.20	028.41	0
028	Manganese (ppm)	0226	104.4	9.460	128.3	7.933	4.158	85	-3.02	028.41	0
028	Manganese (ppm)	0619	111.0	10.00	128.3	7.933	4.158	85	-2.18	028.41	0
028	Manganese (ppm)	2326	112.2	5.200	128.3	7.933	4.158	85	-2.03	028.43	0
028	Manganese (ppm)	0354	116.7	0.3000	128.3	7.933	4.158	85	-1.47	028.41	0
028	Manganese (ppm)	0226	118.0	2.260	128.3	7.933	4.158	85	-1.30	028.43	0
028	Manganese (ppm)	0263	118.8	3.193	128.3	7.933	4.158	85	-1.19	028.42	0
028	Manganese (ppm)	0294	119.5	0.8000	128.3	7.933	4.158	85	-1.11	028.42	0
028	Manganese (ppm)	2319	119.5	3.000	128.3	7.933	4.158	85	-1.11	028.42	0
028	Manganese (ppm)	0511	119.5	1.000	128.3	7.933	4.158	85	-1.11	028.43	0
028	Manganese (ppm)	0674	119.5	0.0000	128.3	7.933	4.158	85	-1.10	028.41	0
028	Manganese (ppm)	0629	120.5	3.000	128.3	7.933	4.158	85	-0.98	028.31	0
028	Manganese (ppm)	0049	120.8	16.95	128.3	7.933	4.158	85	-0.95	028.41	0
028	Manganese (ppm)	2246	121.5	0.3200	128.3	7.933	4.158	85	-0.86	028.31	0
028	Manganese (ppm)	0870	121.6	4.400	128.3	7.933	4.158	85	-0.84	028.42	0
028	Manganese (ppm)	0358	121.7	1.680	128.3	7.933	4.158	85	-0.83	028.41	0
028	Manganese (ppm)	0407	122.0	0.7786	128.3	7.933	4.158	85	-0.79	028.43	0
028	Manganese (ppm)	0242	122.0	10.00	128.3	7.933	4.158	85	-0.79	028.99	0
028	Manganese (ppm)	0921	122.3	4.050	128.3	7.933	4.158	85	-0.76	028.31	0
028	Manganese (ppm)	0598	122.6	2.629	128.3	7.933	4.158	85	-0.71	028.41	0
028	Manganese (ppm)	0553	123.0	4.000	128.3	7.933	4.158	85	-0.67	028.53	0
028	Manganese (ppm)	0123	123.0	0.0000	128.3	7.933	4.158	85	-0.67	028.99	0
028	Manganese (ppm)	2196	123.7	0.0000	128.3	7.933	4.158	85	-0.58	028.31	0
028	Manganese (ppm)	0610	123.8	4.500	128.3	7.933	4.158	85	-0.57	028.43	0
028	Manganese (ppm)	0164	124.0	0.0000	128.3	7.933	4.158	85	-0.54	028.44	0
028	Manganese (ppm)	0407	124.1	5.682	128.3	7.933	4.158	85	-0.53	028.41	0
028	Manganese (ppm)	2129	124.1	0.0000	128.3	7.933	4.158	85	-0.53	028.42	0
028	Manganese (ppm)	0083	124.5	1.910	128.3	7.933	4.158	85	-0.47	028.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
028	Manganese (ppm)	2012	124.8	0.4200	128.3	7.933	4.158	85	-0.44	028.41	0
028	Manganese (ppm)	0512	124.9	4.700	128.3	7.933	4.158	85	-0.43	028.43	0
028	Manganese (ppm)	0169	125.0	4.000	128.3	7.933	4.158	85	-0.41	028.43	0
028	Manganese (ppm)	2089	125.2	0.0500	128.3	7.933	4.158	85	-0.39	028.43	0
028	Manganese (ppm)	0520	125.3	8.021	128.3	7.933	4.158	85	-0.37	028.41	0
028	Manganese (ppm)	0098	125.4	4.200	128.3	7.933	4.158	85	-0.36	028.41	0
028	Manganese (ppm)	0004	125.6	2.380	128.3	7.933	4.158	85	-0.33	028.41	0
028	Manganese (ppm)	2188	125.8	0.0000	128.3	7.933	4.158	85	-0.32	028.31	0
028	Manganese (ppm)	0910	126.0	14.00	128.3	7.933	4.158	85	-0.29	028.41	0
028	Manganese (ppm)	0510	126.0	0.0000	128.3	7.933	4.158	85	-0.29	028.43	0
028	Manganese (ppm)	0870	126.8	6.000	128.3	7.933	4.158	85	-0.19	028.43	0
028	Manganese (ppm)	0693	127.1	0.7700	128.3	7.933	4.158	85	-0.15	028.42	0
028	Manganese (ppm)	0171	127.2	0.1000	128.3	7.933	4.158	85	-0.14	028.41	0
028	Manganese (ppm)	0357	128.0	3.971	128.3	7.933	4.158	85	-0.03	028.42	0
028	Manganese (ppm)	0278	129.0	4.000	128.3	7.933	4.158	85	0.09	028.42	0
028	Manganese (ppm)	2022	129.2	0.9900	128.3	7.933	4.158	85	0.11	028.33	0
028	Manganese (ppm)	0948	129.5	1.830	128.3	7.933	4.158	85	0.16	028.33	0
028	Manganese (ppm)	0939	129.6	2.890	128.3	7.933	4.158	85	0.17	028.31	0
028	Manganese (ppm)	0407	129.7	7.276	128.3	7.933	4.158	85	0.17	028.53	0
028	Manganese (ppm)	0504	129.8	0.1000	128.3	7.933	4.158	85	0.19	028.42	0
028	Manganese (ppm)	0047	130.2	2.550	128.3	7.933	4.158	85	0.24	028.52	0
028	Manganese (ppm)	0651	130.4	0.5210	128.3	7.933	4.158	85	0.27	028.31	0
028	Manganese (ppm)	0033	130.5	13.00	128.3	7.933	4.158	85	0.28	028.43	0
028	Manganese (ppm)	0656	130.7	0.0000	128.3	7.933	4.158	85	0.30	028.31	0
028	Manganese (ppm)	0100	131.5	3.000	128.3	7.933	4.158	85	0.41	028.99	0
028	Manganese (ppm)	0229	131.9	5.700	128.3	7.933	4.158	85	0.45	028.41	0
028	Manganese (ppm)	0035	132.0	0.0000	128.3	7.933	4.158	85	0.47	028.43	0
028	Manganese (ppm)	0148	132.1	2.793	128.3	7.933	4.158	85	0.49	028.41	0
028	Manganese (ppm)	0045	132.5	7.000	128.3	7.933	4.158	85	0.53	028.42	0
028	Manganese (ppm)	0629	132.5	3.000	128.3	7.933	4.158	85	0.53	028.43	0
028	Manganese (ppm)	0626	132.5	3.270	128.3	7.933	4.158	85	0.53	028.43	0
028	Manganese (ppm)	0035	133.0	4.000	128.3	7.933	4.158	85	0.60	028.42	0
028	Manganese (ppm)	2192	133.0	0.0000	128.3	7.933	4.158	85	0.60	028.53	0
028	Manganese (ppm)	0968	133.3	0.6650	128.3	7.933	4.158	85	0.63	028.43	0
028	Manganese (ppm)	0692	133.5	9.000	128.3	7.933	4.158	85	0.66	028.42	0
028	Manganese (ppm)	0042	134.0	4.000	128.3	7.933	4.158	85	0.72	028.42	0
028	Manganese (ppm)	0017	134.2	2.015	128.3	7.933	4.158	85	0.75	028.43	0
028	Manganese (ppm)	0876	134.3	12.50	128.3	7.933	4.158	85	0.75	028.43	0
028	Manganese (ppm)	0682	134.4	0.0000	128.3	7.933	4.158	85	0.77	028.43	0
028	Manganese (ppm)	0027	134.7	4.355	128.3	7.933	4.158	85	0.80	028.43	0
028	Manganese (ppm)	0098	134.8	1.200	128.3	7.933	4.158	85	0.82	028.43	0
028	Manganese (ppm)	0019	135.4	0.6000	128.3	7.933	4.158	85	0.90	028.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
028	Manganese (ppm)	0345	135.8	0.5000	128.3	7.933	4.158	85	0.94	028.43	0
028	Manganese (ppm)	0529	136.2	11.40	128.3	7.933	4.158	85	1.00	028.31	0
028	Manganese (ppm)	0038	136.5	11.00	128.3	7.933	4.158	85	1.04	028.32	0
028	Manganese (ppm)	0675	137.0	1.770	128.3	7.933	4.158	85	1.09	028.43	0
028	Manganese (ppm)	0505	137.9	7.000	128.3	7.933	4.158	85	1.21	028.43	0
028	Manganese (ppm)	0001	139.8	0.9000	128.3	7.933	4.158	85	1.45	028.31	0
028	Manganese (ppm)	0190	139.9	1.420	128.3	7.933	4.158	85	1.47	028.42	0
028	Manganese (ppm)	0009	140.0	8.000	128.3	7.933	4.158	85	1.48	028.42	0
028	Manganese (ppm)	0353	140.5	5.000	128.3	7.933	4.158	85	1.54	028.43	0
028	Manganese (ppm)	0297	141.0	20.00	128.3	7.933	4.158	85	1.60	028.43	0
028	Manganese (ppm)	0202	143.0	16.00	128.3	7.933	4.158	85	1.86	028.43	0
028	Manganese (ppm)	0726	145.0	0.5000	128.3	7.933	4.158	85	2.10	028.42	0
028	Manganese (ppm)	2375	156.5	19.00	128.3	7.933	4.158	85	3.56	028.42	0
028	Manganese (ppm)	0186	156.5	3.000	128.3	7.933	4.158	85	3.56	028.52	0
028	Manganese (ppm)	0265	136.5	29.00	128.3	7.933	4.158	85	1.04	028.42	1
028	Manganese (ppm)	0366	141.5	25.00	128.3	7.933	4.158	85	1.67	028.42	1
030	Nitrate (%)	0357	0.0042	0.0000				1		030.01	0
031	Phosphorus (%)	0108	0.6250	0.0500	0.7054	0.0330	0.0164	121	-2.44	031.01	0
031	Phosphorus (%)	0294	0.6250	0.0100	0.7054	0.0330	0.0164	121	-2.44	031.42	0
031	Phosphorus (%)	0169	0.6300	0.0200	0.7054	0.0330	0.0164	121	-2.29	031.03	0
031	Phosphorus (%)	0018	0.6440	0.0080	0.7054	0.0330	0.0164	121	-1.86	031.01	0
031	Phosphorus (%)	2022	0.6450	0.0100	0.7054	0.0330	0.0164	121	-1.83	031.01	0
031	Phosphorus (%)	2290	0.6450	0.0300	0.7054	0.0330	0.0164	121	-1.83	031.99	0
031	Phosphorus (%)	0066	0.6465	0.0190	0.7054	0.0330	0.0164	121	-1.79	031.44	0
031	Phosphorus (%)	2006	0.6495	0.0290	0.7054	0.0330	0.0164	121	-1.70	031.01	0
031	Phosphorus (%)	0529	0.6550	0.0100	0.7054	0.0330	0.0164	121	-1.53	031.01	0
031	Phosphorus (%)	2200	0.6600	0.0400	0.7054	0.0330	0.0164	121	-1.38	031.01	0
031	Phosphorus (%)	0047	0.6600	0.0000	0.7054	0.0330	0.0164	121	-1.38	031.42	0
031	Phosphorus (%)	0968	0.6625	0.0010	0.7054	0.0330	0.0164	121	-1.30	031.43	0
031	Phosphorus (%)	0019	0.6650	0.0100	0.7054	0.0330	0.0164	121	-1.23	031.41	0
031	Phosphorus (%)	0010	0.6700	0.0000	0.7054	0.0330	0.0164	121	-1.07	031.43	0
031	Phosphorus (%)	0242	0.6700	0.0000	0.7054	0.0330	0.0164	121	-1.07	031.99	0
031	Phosphorus (%)	0939	0.6750	0.0300	0.7054	0.0330	0.0164	121	-0.92	031.01	0
031	Phosphorus (%)	2144	0.6750	0.0100	0.7054	0.0330	0.0164	121	-0.92	031.01	0
031	Phosphorus (%)	0229	0.6750	0.0300	0.7054	0.0330	0.0164	121	-0.92	031.41	0
031	Phosphorus (%)	0265	0.6750	0.0300	0.7054	0.0330	0.0164	121	-0.92	031.42	0
031	Phosphorus (%)	0692	0.6750	0.1500	0.7054	0.0330	0.0164	121	-0.92	031.42	0
031	Phosphorus (%)	0226	0.6750	0.0100	0.7054	0.0330	0.0164	121	-0.92	031.43	0
031	Phosphorus (%)	0676	0.6785	0.0130	0.7054	0.0330	0.0164	121	-0.82	031.99	0
031	Phosphorus (%)	0226	0.6800	0.0200	0.7054	0.0330	0.0164	121	-0.77	031.41	0
031	Phosphorus (%)	0536	0.6800	0.0000	0.7054	0.0330	0.0164	121	-0.77	031.06	0
031	Phosphorus (%)	0674	0.6800	0.0000	0.7054	0.0330	0.0164	121	-0.77	031.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	0407	0.6823	0.0009	0.7054	0.0330	0.0164	121	-0.70	031.43	0
031	Phosphorus (%)	0885	0.6833	0.0056	0.7054	0.0330	0.0164	121	-0.67	031.01	0
031	Phosphorus (%)	2397	0.6850	0.0100	0.7054	0.0330	0.0164	121	-0.62	031.01	0
031	Phosphorus (%)	0910	0.6850	0.0100	0.7054	0.0330	0.0164	121	-0.62	031.41	0
031	Phosphorus (%)	0033	0.6850	0.0100	0.7054	0.0330	0.0164	121	-0.62	031.43	0
031	Phosphorus (%)	0511	0.6850	0.0500	0.7054	0.0330	0.0164	121	-0.62	031.43	0
031	Phosphorus (%)	0610	0.6850	0.0100	0.7054	0.0330	0.0164	121	-0.62	031.43	0
031	Phosphorus (%)	2129	0.6852	0.0057	0.7054	0.0330	0.0164	121	-0.61	031.42	0
031	Phosphorus (%)	0038	0.6860	0.0080	0.7054	0.0330	0.0164	121	-0.59	031.01	0
031	Phosphorus (%)	0619	0.6860	0.0040	0.7054	0.0330	0.0164	121	-0.59	031.41	0
031	Phosphorus (%)	0009	0.6866	0.0057	0.7054	0.0330	0.0164	121	-0.57	031.42	0
031	Phosphorus (%)	0512	0.6887	0.0088	0.7054	0.0330	0.0164	121	-0.51	031.43	0
031	Phosphorus (%)	2109	0.6890	0.0180	0.7054	0.0330	0.0164	121	-0.50	031.01	0
031	Phosphorus (%)	2192	0.6900	0.0200	0.7054	0.0330	0.0164	121	-0.47	031.01	0
031	Phosphorus (%)	0693	0.6900	0.0080	0.7054	0.0330	0.0164	121	-0.47	031.42	0
031	Phosphorus (%)	0297	0.6900	0.0400	0.7054	0.0330	0.0164	121	-0.47	031.43	0
031	Phosphorus (%)	0100	0.6900	0.0000	0.7054	0.0330	0.0164	121	-0.47	031.99	0
031	Phosphorus (%)	0083	0.6925	0.0030	0.7054	0.0330	0.0164	121	-0.39	031.43	0
031	Phosphorus (%)	0142	0.6929	0.0146	0.7054	0.0330	0.0164	121	-0.38	031.01	0
031	Phosphorus (%)	0045	0.6935	0.0630	0.7054	0.0330	0.0164	121	-0.36	031.42	0
031	Phosphorus (%)	2319	0.6935	0.0050	0.7054	0.0330	0.0164	121	-0.36	031.42	0
031	Phosphorus (%)	0629	0.6950	0.0100	0.7054	0.0330	0.0164	121	-0.32	031.01	0
031	Phosphorus (%)	2146	0.6950	0.0100	0.7054	0.0330	0.0164	121	-0.32	031.01	0
031	Phosphorus (%)	0004	0.6950	0.0100	0.7054	0.0330	0.0164	121	-0.32	031.41	0
031	Phosphorus (%)	0848	0.6950	0.0100	0.7054	0.0330	0.0164	121	-0.32	031.41	0
031	Phosphorus (%)	2340	0.6960	0.0020	0.7054	0.0330	0.0164	121	-0.29	031.01	0
031	Phosphorus (%)	2314	0.6965	0.0023	0.7054	0.0330	0.0164	121	-0.27	031.43	0
031	Phosphorus (%)	0651	0.6969	0.0105	0.7054	0.0330	0.0164	121	-0.26	031.01	0
031	Phosphorus (%)	2199	0.7000	0.0200	0.7054	0.0330	0.0164	121	-0.16	031.01	0
031	Phosphorus (%)	0042	0.7000	0.0100	0.7054	0.0330	0.0164	121	-0.16	031.42	0
031	Phosphorus (%)	0263	0.7000	0.0140	0.7054	0.0330	0.0164	121	-0.16	031.42	0
031	Phosphorus (%)	0089	0.7000	0.0000	0.7054	0.0330	0.0164	121	-0.16	031.43	0
031	Phosphorus (%)	2190	0.7007	0.0182	0.7054	0.0330	0.0164	121	-0.14	031.01	0
031	Phosphorus (%)	2009	0.7008	0.0019	0.7054	0.0330	0.0164	121	-0.14	031.01	0
031	Phosphorus (%)	0598	0.7038	0.0009	0.7054	0.0330	0.0164	121	-0.05	031.41	0
031	Phosphorus (%)	0001	0.7040	0.0160	0.7054	0.0330	0.0164	121	-0.04	031.03	0
031	Phosphorus (%)	0683	0.7050	0.0100	0.7054	0.0330	0.0164	121	-0.01	031.01	0
031	Phosphorus (%)	2325	0.7050	0.0100	0.7054	0.0330	0.0164	121	-0.01	031.01	0
031	Phosphorus (%)	0144	0.7050	0.0300	0.7054	0.0330	0.0164	121	-0.01	031.41	0
031	Phosphorus (%)	0629	0.7050	0.0100	0.7054	0.0330	0.0164	121	-0.01	031.43	0
031	Phosphorus (%)	0164	0.7055	0.0050	0.7054	0.0330	0.0164	121	0.00	031.44	0
031	Phosphorus (%)	0001	0.7085	0.0110	0.7054	0.0330	0.0164	121	0.09	031.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	0006	0.7085	0.0050	0.7054	0.0330	0.0164	121	0.09	031.43	0
031	Phosphorus (%)	0687	0.7100	0.0000	0.7054	0.0330	0.0164	121	0.14	031.01	0
031	Phosphorus (%)	0948	0.7100	0.0200	0.7054	0.0330	0.0164	121	0.14	031.01	0
031	Phosphorus (%)	2246	0.7100	0.0000	0.7054	0.0330	0.0164	121	0.14	031.01	0
031	Phosphorus (%)	2012	0.7100	0.0200	0.7054	0.0330	0.0164	121	0.14	031.41	0
031	Phosphorus (%)	0682	0.7100	0.0000	0.7054	0.0330	0.0164	121	0.14	031.43	0
031	Phosphorus (%)	0123	0.7100	0.0200	0.7054	0.0330	0.0164	121	0.14	031.99	0
031	Phosphorus (%)	2259	0.7125	0.0010	0.7054	0.0330	0.0164	121	0.21	031.41	0
031	Phosphorus (%)	0354	0.7145	0.0010	0.7054	0.0330	0.0164	121	0.28	031.41	0
031	Phosphorus (%)	2302	0.7150	0.0100	0.7054	0.0330	0.0164	121	0.29	031.01	0
031	Phosphorus (%)	0358	0.7150	0.0300	0.7054	0.0330	0.0164	121	0.29	031.41	0
031	Phosphorus (%)	0675	0.7150	0.0100	0.7054	0.0330	0.0164	121	0.29	031.43	0
031	Phosphorus (%)	2146	0.7150	0.0100	0.7054	0.0330	0.0164	121	0.29	031.51	0
031	Phosphorus (%)	2196	0.7160	0.0000	0.7054	0.0330	0.0164	121	0.32	031.01	0
031	Phosphorus (%)	2188	0.7200	0.0000	0.7054	0.0330	0.0164	121	0.44	031.01	0
031	Phosphorus (%)	2337	0.7200	0.0000	0.7054	0.0330	0.0164	121	0.44	031.01	0
031	Phosphorus (%)	0035	0.7200	0.0000	0.7054	0.0330	0.0164	121	0.44	031.42	0
031	Phosphorus (%)	0027	0.7205	0.0050	0.7054	0.0330	0.0164	121	0.46	031.43	0
031	Phosphorus (%)	0723	0.7221	0.0011	0.7054	0.0330	0.0164	121	0.50	031.41	0
031	Phosphorus (%)	0345	0.7225	0.0050	0.7054	0.0330	0.0164	121	0.52	031.43	0
031	Phosphorus (%)	0148	0.7240	0.0100	0.7054	0.0330	0.0164	121	0.56	031.41	0
031	Phosphorus (%)	0202	0.7250	0.0560	0.7054	0.0330	0.0164	121	0.59	031.43	0
031	Phosphorus (%)	2089	0.7250	0.0100	0.7054	0.0330	0.0164	121	0.59	031.43	0
031	Phosphorus (%)	0407	0.7260	0.0440	0.7054	0.0330	0.0164	121	0.62	031.41	0
031	Phosphorus (%)	0357	0.7273	0.0212	0.7054	0.0330	0.0164	121	0.66	031.42	0
031	Phosphorus (%)	0017	0.7300	0.0113	0.7054	0.0330	0.0164	121	0.74	031.43	0
031	Phosphorus (%)	0407	0.7300	0.0311	0.7054	0.0330	0.0164	121	0.74	031.53	0
031	Phosphorus (%)	0049	0.7300	0.0600	0.7054	0.0330	0.0164	121	0.75	031.41	0
031	Phosphorus (%)	0366	0.7300	0.0200	0.7054	0.0330	0.0164	121	0.75	031.42	0
031	Phosphorus (%)	0504	0.7310	0.0400	0.7054	0.0330	0.0164	121	0.78	031.42	0
031	Phosphorus (%)	0553	0.7320	0.0140	0.7054	0.0330	0.0164	121	0.81	031.53	0
031	Phosphorus (%)	0171	0.7350	0.0100	0.7054	0.0330	0.0164	121	0.90	031.41	0
031	Phosphorus (%)	0505	0.7350	0.0100	0.7054	0.0330	0.0164	121	0.90	031.43	0
031	Phosphorus (%)	0626	0.7355	0.0130	0.7054	0.0330	0.0164	121	0.91	031.43	0
031	Phosphorus (%)	0098	0.7379	0.0059	0.7054	0.0330	0.0164	121	0.98	031.41	0
031	Phosphorus (%)	0035	0.7400	0.0200	0.7054	0.0330	0.0164	121	1.05	031.43	0
031	Phosphorus (%)	0300	0.7400	0.0200	0.7054	0.0330	0.0164	121	1.05	031.43	0
031	Phosphorus (%)	0014	0.7415	0.0150	0.7054	0.0330	0.0164	121	1.09	031.42	0
031	Phosphorus (%)	0689	0.7500	0.0200	0.7054	0.0330	0.0164	121	1.35	031.01	0
031	Phosphorus (%)	0190	0.7500	0.0000	0.7054	0.0330	0.0164	121	1.35	031.42	0
031	Phosphorus (%)	0036	0.7545	0.0206	0.7054	0.0330	0.0164	121	1.49	031.43	0
031	Phosphorus (%)	0510	0.7600	0.0400	0.7054	0.0330	0.0164	121	1.65	031.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
031	Phosphorus (%)	0098	0.7618	0.0121	0.7054	0.0330	0.0164	121	1.71	031.43	0
031	Phosphorus (%)	0726	0.7650	0.0100	0.7054	0.0330	0.0164	121	1.81	031.42	0
031	Phosphorus (%)	0876	0.7658	0.0115	0.7054	0.0330	0.0164	121	1.83	031.43	0
031	Phosphorus (%)	0353	0.7750	0.0100	0.7054	0.0330	0.0164	121	2.11	031.43	0
031	Phosphorus (%)	0186	0.7769	0.0020	0.7054	0.0330	0.0164	121	2.17	031.52	0
031	Phosphorus (%)	0656	0.7800	0.0000	0.7054	0.0330	0.0164	121	2.26	031.01	0
031	Phosphorus (%)	0870	0.7816	0.0452	0.7054	0.0330	0.0164	121	2.31	031.42	0
031	Phosphorus (%)	0278	0.7900	0.0400	0.7054	0.0330	0.0164	121	2.56	031.42	0
031	Phosphorus (%)	0520	0.8090	0.0340	0.7054	0.0330	0.0164	121	3.14	031.41	0
031	Phosphorus (%)	0870	0.8521	0.0076	0.7054	0.0330	0.0164	121	4.45	031.43	0
031	Phosphorus (%)	2375	0.9050	0.0900	0.7054	0.0330	0.0164	121	6.05	031.42	0
031	Phosphorus (%)	0596	1.035	0.0500	0.7054	0.0330	0.0164	121	9.99	031.01	0
031	Phosphorus (%)	2326	1.060	0.4800	0.7054	0.0330	0.0164	121	10.75	031.43	2
032	Potassium (%)	2326	0.7850	0.0700	1.040	0.0528	0.0252	87	-4.84	032.43	0
032	Potassium (%)	0226	0.8000	0.0000	1.040	0.0528	0.0252	87	-4.55	032.43	0
032	Potassium (%)	0169	0.8350	0.0100	1.040	0.0528	0.0252	87	-3.89	032.32	0
032	Potassium (%)	0004	0.8950	0.1500	1.040	0.0528	0.0252	87	-2.75	032.41	0
032	Potassium (%)	0038	0.9085	0.0330	1.040	0.0528	0.0252	87	-2.50	032.31	0
032	Potassium (%)	0910	0.9300	0.0800	1.040	0.0528	0.0252	87	-2.09	032.41	0
032	Potassium (%)	0019	0.9450	0.0100	1.040	0.0528	0.0252	87	-1.80	032.41	0
032	Potassium (%)	2129	0.9476	0.0026	1.040	0.0528	0.0252	87	-1.76	032.42	0
032	Potassium (%)	0089	0.9500	0.0000	1.040	0.0528	0.0252	87	-1.71	032.43	0
032	Potassium (%)	0683	0.9650	0.0100	1.040	0.0528	0.0252	87	-1.43	032.31	0
032	Potassium (%)	0619	0.9665	0.0050	1.040	0.0528	0.0252	87	-1.40	032.41	0
032	Potassium (%)	0294	0.9850	0.0100	1.040	0.0528	0.0252	87	-1.05	032.42	0
032	Potassium (%)	0512	0.9898	0.0665	1.040	0.0528	0.0252	87	-0.96	032.43	0
032	Potassium (%)	0229	0.9900	0.0000	1.040	0.0528	0.0252	87	-0.95	032.41	0
032	Potassium (%)	0674	0.9900	0.0000	1.040	0.0528	0.0252	87	-0.95	032.41	0
032	Potassium (%)	2012	0.9950	0.0300	1.040	0.0528	0.0252	87	-0.86	032.41	0
032	Potassium (%)	0870	0.9989	0.0002	1.040	0.0528	0.0252	87	-0.78	032.42	0
032	Potassium (%)	0629	1.005	0.0300	1.040	0.0528	0.0252	87	-0.67	032.31	0
032	Potassium (%)	0033	1.005	0.0100	1.040	0.0528	0.0252	87	-0.67	032.43	0
032	Potassium (%)	0511	1.005	0.0100	1.040	0.0528	0.0252	87	-0.67	032.43	0
032	Potassium (%)	0242	1.010	0.0000	1.040	0.0528	0.0252	87	-0.57	032.99	0
032	Potassium (%)	0870	1.013	0.0295	1.040	0.0528	0.0252	87	-0.51	032.43	0
032	Potassium (%)	0066	1.014	0.0420	1.040	0.0528	0.0252	87	-0.50	032.41	0
032	Potassium (%)	0968	1.015	0.0210	1.040	0.0528	0.0252	87	-0.49	032.43	0
032	Potassium (%)	0529	1.015	0.0100	1.040	0.0528	0.0252	87	-0.48	032.31	0
032	Potassium (%)	0098	1.015	0.0500	1.040	0.0528	0.0252	87	-0.48	032.43	0
032	Potassium (%)	0629	1.015	0.0100	1.040	0.0528	0.0252	87	-0.48	032.43	0
032	Potassium (%)	2089	1.015	0.0100	1.040	0.0528	0.0252	87	-0.48	032.43	0
032	Potassium (%)	0407	1.019	0.0143	1.040	0.0528	0.0252	87	-0.40	032.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
032	Potassium (%)	0049	1.020	0.0800	1.040	0.0528	0.0252	87	-0.38	032.41	0
032	Potassium (%)	0144	1.020	0.0000	1.040	0.0528	0.0252	87	-0.38	032.41	0
032	Potassium (%)	0142	1.021	0.0277	1.040	0.0528	0.0252	87	-0.37	032.31	0
032	Potassium (%)	2319	1.025	0.0100	1.040	0.0528	0.0252	87	-0.29	032.52	0
032	Potassium (%)	0553	1.025	0.0300	1.040	0.0528	0.0252	87	-0.29	032.53	0
032	Potassium (%)	0598	1.029	0.0004	1.040	0.0528	0.0252	87	-0.21	032.41	0
032	Potassium (%)	0939	1.030	0.0200	1.040	0.0528	0.0252	87	-0.19	032.31	0
032	Potassium (%)	0164	1.030	0.0000	1.040	0.0528	0.0252	87	-0.19	032.44	0
032	Potassium (%)	0100	1.030	0.0200	1.040	0.0528	0.0252	87	-0.19	032.99	0
032	Potassium (%)	0520	1.031	0.0060	1.040	0.0528	0.0252	87	-0.18	032.41	0
032	Potassium (%)	0226	1.035	0.0100	1.040	0.0528	0.0252	87	-0.10	032.41	0
032	Potassium (%)	0035	1.035	0.0100	1.040	0.0528	0.0252	87	-0.10	032.42	0
032	Potassium (%)	0047	1.040	0.0200	1.040	0.0528	0.0252	87	-0.01	032.42	0
032	Potassium (%)	0265	1.040	0.0800	1.040	0.0528	0.0252	87	-0.01	032.42	0
032	Potassium (%)	0610	1.040	0.0000	1.040	0.0528	0.0252	87	-0.01	032.43	0
032	Potassium (%)	2394	1.040	0.0000	1.040	0.0528	0.0252	87	-0.01	032.43	0
032	Potassium (%)	0407	1.042	0.0307	1.040	0.0528	0.0252	87	0.03	032.53	0
032	Potassium (%)	0345	1.045	0.0100	1.040	0.0528	0.0252	87	0.09	032.43	0
032	Potassium (%)	0407	1.045	0.0488	1.040	0.0528	0.0252	87	0.10	032.41	0
032	Potassium (%)	0357	1.046	0.0028	1.040	0.0528	0.0252	87	0.12	032.42	0
032	Potassium (%)	0083	1.050	0.0000	1.040	0.0528	0.0252	87	0.18	032.43	0
032	Potassium (%)	0027	1.053	0.0180	1.040	0.0528	0.0252	87	0.24	032.43	0
032	Potassium (%)	0171	1.055	0.0100	1.040	0.0528	0.0252	87	0.28	032.41	0
032	Potassium (%)	0693	1.055	0.0500	1.040	0.0528	0.0252	87	0.28	032.42	0
032	Potassium (%)	0148	1.060	0.0200	1.040	0.0528	0.0252	87	0.37	032.41	0
032	Potassium (%)	0358	1.060	0.0200	1.040	0.0528	0.0252	87	0.37	032.41	0
032	Potassium (%)	0042	1.060	0.0200	1.040	0.0528	0.0252	87	0.37	032.42	0
032	Potassium (%)	0006	1.060	0.0000	1.040	0.0528	0.0252	87	0.37	032.43	0
032	Potassium (%)	0510	1.060	0.0200	1.040	0.0528	0.0252	87	0.37	032.43	0
032	Potassium (%)	0123	1.060	0.0000	1.040	0.0528	0.0252	87	0.37	032.99	0
032	Potassium (%)	0504	1.065	0.0470	1.040	0.0528	0.0252	87	0.46	032.42	0
032	Potassium (%)	0692	1.065	0.1500	1.040	0.0528	0.0252	87	0.47	032.42	0
032	Potassium (%)	0035	1.065	0.0100	1.040	0.0528	0.0252	87	0.47	032.43	0
032	Potassium (%)	0036	1.065	0.0260	1.040	0.0528	0.0252	87	0.47	032.43	0
032	Potassium (%)	0948	1.067	0.0080	1.040	0.0528	0.0252	87	0.51	032.31	0
032	Potassium (%)	0263	1.069	0.0280	1.040	0.0528	0.0252	87	0.54	032.42	0
032	Potassium (%)	0354	1.070	0.0020	1.040	0.0528	0.0252	87	0.56	032.41	0
032	Potassium (%)	2246	1.070	0.0000	1.040	0.0528	0.0252	87	0.56	032.31	0
032	Potassium (%)	0098	1.070	0.0000	1.040	0.0528	0.0252	87	0.56	032.41	0
032	Potassium (%)	0297	1.075	0.0900	1.040	0.0528	0.0252	87	0.66	032.43	0
032	Potassium (%)	0682	1.080	0.0000	1.040	0.0528	0.0252	87	0.75	032.43	0
032	Potassium (%)	0876	1.080	0.0200	1.040	0.0528	0.0252	87	0.75	032.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
032	Potassium (%)	0186	1.081	0.0093	1.040	0.0528	0.0252	87	0.78	032.42	0
032	Potassium (%)	0017	1.089	0.0015	1.040	0.0528	0.0252	87	0.93	032.43	0
032	Potassium (%)	0202	1.090	0.0600	1.040	0.0528	0.0252	87	0.94	032.43	0
032	Potassium (%)	0010	1.095	0.0100	1.040	0.0528	0.0252	87	1.04	032.43	0
032	Potassium (%)	0626	1.102	0.0240	1.040	0.0528	0.0252	87	1.17	032.43	0
032	Potassium (%)	0001	1.107	0.0190	1.040	0.0528	0.0252	87	1.25	032.99	0
032	Potassium (%)	0278	1.110	0.0200	1.040	0.0528	0.0252	87	1.32	032.42	0
032	Potassium (%)	0675	1.115	0.0100	1.040	0.0528	0.0252	87	1.42	032.43	0
032	Potassium (%)	0009	1.123	0.0475	1.040	0.0528	0.0252	87	1.56	032.42	0
032	Potassium (%)	0190	1.130	0.0000	1.040	0.0528	0.0252	87	1.70	032.42	0
032	Potassium (%)	0186	1.135	0.0053	1.040	0.0528	0.0252	87	1.80	032.52	0
032	Potassium (%)	0353	1.145	0.0300	1.040	0.0528	0.0252	87	1.98	032.43	0
032	Potassium (%)	0726	1.150	0.0400	1.040	0.0528	0.0252	87	2.08	032.42	0
032	Potassium (%)	0505	1.180	0.0800	1.040	0.0528	0.0252	87	2.65	032.43	0
032	Potassium (%)	0366	1.220	0.0600	1.040	0.0528	0.0252	87	3.40	032.42	0
032	Potassium (%)	2375	1.310	0.1200	1.040	0.0528	0.0252	87	5.11	032.42	0
032	Potassium (%)	0045	0.9265	0.2470	1.040	0.0528	0.0252	87	-2.16	032.42	1
033	Salt as chloride (%)	0265	0.6200	0.0000	0.8034	0.0410	0.0131	63	-4.48	033.03	0
033	Salt as chloride (%)	2337	0.6300	0.0000	0.8034	0.0410	0.0131	63	-4.23	033.00	0
033	Salt as chloride (%)	0171	0.6700	0.0000	0.8034	0.0410	0.0131	63	-3.26	033.05	0
033	Salt as chloride (%)	0619	0.6980	0.0160	0.8034	0.0410	0.0131	63	-2.57	033.01	0
033	Salt as chloride (%)	0674	0.7000	0.0000	0.8034	0.0410	0.0131	63	-2.52	033.00	0
033	Salt as chloride (%)	2073	0.7550	0.0100	0.8034	0.0410	0.0131	63	-1.18	033.99	0
033	Salt as chloride (%)	0123	0.7600	0.0000	0.8034	0.0410	0.0131	63	-1.06	033.00	0
033	Salt as chloride (%)	2196	0.7600	0.0000	0.8034	0.0410	0.0131	63	-1.06	033.00	0
033	Salt as chloride (%)	0910	0.7600	0.0200	0.8034	0.0410	0.0131	63	-1.06	033.99	0
033	Salt as chloride (%)	0689	0.7650	0.0300	0.8034	0.0410	0.0131	63	-0.94	033.05	0
033	Salt as chloride (%)	2009	0.7668	0.0130	0.8034	0.0410	0.0131	63	-0.89	033.00	0
033	Salt as chloride (%)	0278	0.7750	0.0100	0.8034	0.0410	0.0131	63	-0.69	033.01	0
033	Salt as chloride (%)	0510	0.7750	0.0000	0.8034	0.0410	0.0131	63	-0.69	033.01	0
033	Salt as chloride (%)	2109	0.7780	0.0060	0.8034	0.0410	0.0131	63	-0.62	033.01	0
033	Salt as chloride (%)	0309	0.7796	0.0171	0.8034	0.0410	0.0131	63	-0.58	033.00	0
033	Salt as chloride (%)	0190	0.7800	0.0000	0.8034	0.0410	0.0131	63	-0.57	033.03	0
033	Salt as chloride (%)	0726	0.7800	0.0000	0.8034	0.0410	0.0131	63	-0.57	033.03	0
033	Salt as chloride (%)	0229	0.7850	0.0100	0.8034	0.0410	0.0131	63	-0.45	033.01	0
033	Salt as chloride (%)	0948	0.7850	0.0100	0.8034	0.0410	0.0131	63	-0.45	033.01	0
033	Salt as chloride (%)	0354	0.7890	0.0020	0.8034	0.0410	0.0131	63	-0.35	033.01	0
033	Salt as chloride (%)	0353	0.7900	0.0200	0.8034	0.0410	0.0131	63	-0.33	033.00	0
033	Salt as chloride (%)	0683	0.7900	0.0000	0.8034	0.0410	0.0131	63	-0.33	033.01	0
033	Salt as chloride (%)	0505	0.7900	0.0600	0.8034	0.0410	0.0131	63	-0.33	033.03	0
033	Salt as chloride (%)	0650	0.7900	0.0000	0.8034	0.0410	0.0131	63	-0.33	033.99	0
033	Salt as chloride (%)	2190	0.7916	0.0052	0.8034	0.0410	0.0131	63	-0.29	033.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
033	Salt as chloride (%)	0148	0.7920	0.0020	0.8034	0.0410	0.0131	63	-0.28	033.01	0
033	Salt as chloride (%)	2319	0.7925	0.0070	0.8034	0.0410	0.0131	63	-0.27	033.01	0
033	Salt as chloride (%)	2193	0.7950	0.0100	0.8034	0.0410	0.0131	63	-0.21	033.00	0
033	Salt as chloride (%)	0001	0.7950	0.0100	0.8034	0.0410	0.0131	63	-0.21	033.01	0
033	Salt as chloride (%)	0164	0.7950	0.0100	0.8034	0.0410	0.0131	63	-0.21	033.01	0
033	Salt as chloride (%)	2146	0.7950	0.0100	0.8034	0.0410	0.0131	63	-0.21	033.01	0
033	Salt as chloride (%)	0693	0.8000	0.0200	0.8034	0.0410	0.0131	63	-0.08	033.00	0
033	Salt as chloride (%)	0003	0.8000	0.0000	0.8034	0.0410	0.0131	63	-0.08	033.01	0
033	Salt as chloride (%)	0028	0.8000	0.0000	0.8034	0.0410	0.0131	63	-0.08	033.01	0
033	Salt as chloride (%)	0083	0.8000	0.0000	0.8034	0.0410	0.0131	63	-0.08	033.01	0
033	Salt as chloride (%)	0226	0.8000	0.0000	0.8034	0.0410	0.0131	63	-0.08	033.01	0
033	Salt as chloride (%)	2375	0.8000	0.0000	0.8034	0.0410	0.0131	63	-0.08	033.01	0
033	Salt as chloride (%)	2022	0.8050	0.0300	0.8034	0.0410	0.0131	63	0.04	033.00	0
033	Salt as chloride (%)	0098	0.8050	0.0100	0.8034	0.0410	0.0131	63	0.04	033.01	0
033	Salt as chloride (%)	0407	0.8050	0.0080	0.8034	0.0410	0.0131	63	0.04	033.01	0
033	Salt as chloride (%)	0004	0.8100	0.0200	0.8034	0.0410	0.0131	63	0.16	033.01	0
033	Salt as chloride (%)	0242	0.8100	0.0000	0.8034	0.0410	0.0131	63	0.16	033.01	0
033	Salt as chloride (%)	0939	0.8150	0.0100	0.8034	0.0410	0.0131	63	0.28	033.01	0
033	Salt as chloride (%)	2326	0.8200	0.0000	0.8034	0.0410	0.0131	63	0.40	033.99	0
033	Salt as chloride (%)	0885	0.8220	0.0338	0.8034	0.0410	0.0131	63	0.45	033.00	0
033	Salt as chloride (%)	0045	0.8250	0.0300	0.8034	0.0410	0.0131	63	0.53	033.00	0
033	Salt as chloride (%)	0539	0.8250	0.0100	0.8034	0.0410	0.0131	63	0.53	033.00	0
033	Salt as chloride (%)	0675	0.8300	0.0200	0.8034	0.0410	0.0131	63	0.65	033.00	0
033	Salt as chloride (%)	0100	0.8300	0.0000	0.8034	0.0410	0.0131	63	0.65	033.01	0
033	Salt as chloride (%)	2268	0.8300	0.0200	0.8034	0.0410	0.0131	63	0.65	033.01	0
033	Salt as chloride (%)	0019	0.8350	0.0100	0.8034	0.0410	0.0131	63	0.77	033.01	0
033	Salt as chloride (%)	2129	0.8394	0.0079	0.8034	0.0410	0.0131	63	0.88	033.99	0
033	Salt as chloride (%)	0511	0.8500	0.0600	0.8034	0.0410	0.0131	63	1.14	033.00	0
033	Salt as chloride (%)	0682	0.8600	0.0000	0.8034	0.0410	0.0131	63	1.38	033.01	0
033	Salt as chloride (%)	0144	0.8600	0.0000	0.8034	0.0410	0.0131	63	1.38	033.03	0
033	Salt as chloride (%)	2076	0.8665	0.0410	0.8034	0.0410	0.0131	63	1.54	033.00	0
033	Salt as chloride (%)	0358	0.9400	0.0400	0.8034	0.0410	0.0131	63	3.33	033.99	0
033	Salt as chloride (%)	0300	0.9430	0.0440	0.8034	0.0410	0.0131	63	3.41	033.99	0
033	Salt as chloride (%)	0017	0.9549	0.0088	0.8034	0.0410	0.0131	63	3.70	033.99	0
033	Salt as chloride (%)	0027	0.9675	0.0150	0.8034	0.0410	0.0131	63	4.00	033.99	0
033	Salt as chloride (%)	0038	0.9795	0.0610	0.8034	0.0410	0.0131	63	4.30	033.01	0
033	Salt as chloride (%)	2397	1.005	0.0500	0.8034	0.0410	0.0131	63	4.92	033.00	0
033	Salt as chloride (%)	2188	1.240	0.0000	0.8034	0.0410	0.0131	63	10.65	033.05	0
033	Salt as chloride (%)	0366	0.5050	0.0700	0.8034	0.0410	0.0131	63	-7.28	033.00	1
034	Selenium, Total (Se) (ppm)	0619	0.3580	0.0080	0.5954	0.1425	0.0328	19	-1.67	034.41	0
034	Selenium, Total (Se) (ppm)	0047	0.4280	0.0160	0.5954	0.1425	0.0328	19	-1.17	034.52	0
034	Selenium, Total (Se) (ppm)	0066	0.4610	0.0240	0.5954	0.1425	0.0328	19	-0.94	034.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
034	Selenium, Total (Se) (ppm)	0009	0.4738	0.0263	0.5954	0.1425	0.0328	19	-0.85	034.53	0
034	Selenium, Total (Se) (ppm)	0512	0.4904	0.0079	0.5954	0.1425	0.0328	19	-0.74	034.04	0
034	Selenium, Total (Se) (ppm)	0045	0.5130	0.1040	0.5954	0.1425	0.0328	19	-0.58	034.04	0
034	Selenium, Total (Se) (ppm)	0553	0.5245	0.0290	0.5954	0.1425	0.0328	19	-0.50	034.53	0
034	Selenium, Total (Se) (ppm)	0003	0.5885	0.0950	0.5954	0.1425	0.0328	19	-0.05	034.53	0
034	Selenium, Total (Se) (ppm)	0164	0.5925	0.0530	0.5954	0.1425	0.0328	19	-0.02	034.53	0
034	Selenium, Total (Se) (ppm)	0560	0.6004	0.0063	0.5954	0.1425	0.0328	19	0.03	034.53	0
034	Selenium, Total (Se) (ppm)	0169	0.6100	0.0400	0.5954	0.1425	0.0328	19	0.10	034.04	0
034	Selenium, Total (Se) (ppm)	2012	0.6140	0.0320	0.5954	0.1425	0.0328	19	0.13	034.53	0
034	Selenium, Total (Se) (ppm)	0610	0.6200	0.0200	0.5954	0.1425	0.0328	19	0.17	034.53	0
034	Selenium, Total (Se) (ppm)	0148	0.6275	0.0430	0.5954	0.1425	0.0328	19	0.23	034.41	0
034	Selenium, Total (Se) (ppm)	0038	0.6560	0.0040	0.5954	0.1425	0.0328	19	0.43	034.01	0
034	Selenium, Total (Se) (ppm)	0407	0.7038	0.0120	0.5954	0.1425	0.0328	19	0.76	034.53	0
034	Selenium, Total (Se) (ppm)	0186	0.9775	0.0130	0.5954	0.1425	0.0328	19	2.68	034.52	0
034	Selenium, Total (Se) (ppm)	0505	1.580	0.0800	0.5954	0.1425	0.0328	19	6.91	034.43	0
034	Selenium, Total (Se) (ppm)	0190	1.625	0.0100	0.5954	0.1425	0.0328	19	7.23	034.99	0
034	Selenium, Total (Se) (ppm)	0692	0.0000	0.0000	0.5954	0.1425	0.0328	19		034.42	4
034	Selenium, Total (Se) (ppm)	0682	< 5		0.5954	0.1425	0.0328	19		034.43	5
035	Sodium (%)	2109	0.2060	0.0220	0.3662	0.0197	0.0098	85	-8.11	035.01	0
035	Sodium (%)	0169	0.2750	0.0100	0.3662	0.0197	0.0098	85	-4.62	035.32	0
035	Sodium (%)	2326	0.2800	0.0200	0.3662	0.0197	0.0098	85	-4.36	035.43	0
035	Sodium (%)	0226	0.3100	0.0000	0.3662	0.0197	0.0098	85	-2.84	035.43	0
035	Sodium (%)	0683	0.3250	0.0100	0.3662	0.0197	0.0098	85	-2.08	035.31	0
035	Sodium (%)	0066	0.3305	0.0370	0.3662	0.0197	0.0098	85	-1.81	035.41	0
035	Sodium (%)	0004	0.3350	0.0300	0.3662	0.0197	0.0098	85	-1.58	035.41	0
035	Sodium (%)	0511	0.3350	0.0100	0.3662	0.0197	0.0098	85	-1.58	035.43	0
035	Sodium (%)	0512	0.3366	0.0127	0.3662	0.0197	0.0098	85	-1.50	035.43	0
035	Sodium (%)	0407	0.3417	0.0531	0.3662	0.0197	0.0098	85	-1.24	035.41	0
035	Sodium (%)	2129	0.3422	0.0029	0.3662	0.0197	0.0098	85	-1.22	035.42	0
035	Sodium (%)	0229	0.3450	0.0100	0.3662	0.0197	0.0098	85	-1.07	035.41	0
035	Sodium (%)	0619	0.3470	0.0020	0.3662	0.0197	0.0098	85	-0.97	035.41	0
035	Sodium (%)	0407	0.3485	0.0017	0.3662	0.0197	0.0098	85	-0.90	035.43	0
035	Sodium (%)	0674	0.3500	0.0000	0.3662	0.0197	0.0098	85	-0.82	035.41	0
035	Sodium (%)	0910	0.3500	0.0000	0.3662	0.0197	0.0098	85	-0.82	035.41	0
035	Sodium (%)	0629	0.3500	0.0000	0.3662	0.0197	0.0098	85	-0.82	035.43	0
035	Sodium (%)	0345	0.3505	0.0010	0.3662	0.0197	0.0098	85	-0.79	035.43	0
035	Sodium (%)	0357	0.3510	0.0035	0.3662	0.0197	0.0098	85	-0.77	035.42	0
035	Sodium (%)	0520	0.3524	0.0031	0.3662	0.0197	0.0098	85	-0.70	035.41	0
035	Sodium (%)	0553	0.3535	0.0070	0.3662	0.0197	0.0098	85	-0.64	035.53	0
035	Sodium (%)	0629	0.3550	0.0100	0.3662	0.0197	0.0098	85	-0.56	035.31	0
035	Sodium (%)	0939	0.3550	0.0100	0.3662	0.0197	0.0098	85	-0.56	035.31	0
035	Sodium (%)	0144	0.3550	0.0100	0.3662	0.0197	0.0098	85	-0.56	035.41	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
035	Sodium (%)	0100	0.3550	0.0300	0.3662	0.0197	0.0098	85	-0.56	035.99	0
035	Sodium (%)	2319	0.3575	0.0030	0.3662	0.0197	0.0098	85	-0.44	035.42	0
035	Sodium (%)	0164	0.3595	0.0010	0.3662	0.0197	0.0098	85	-0.34	035.41	0
035	Sodium (%)	2196	0.3600	0.0000	0.3662	0.0197	0.0098	85	-0.31	035.31	0
035	Sodium (%)	0049	0.3600	0.0400	0.3662	0.0197	0.0098	85	-0.31	035.41	0
035	Sodium (%)	0171	0.3600	0.0000	0.3662	0.0197	0.0098	85	-0.31	035.41	0
035	Sodium (%)	0033	0.3600	0.0200	0.3662	0.0197	0.0098	85	-0.31	035.43	0
035	Sodium (%)	0089	0.3600	0.0000	0.3662	0.0197	0.0098	85	-0.31	035.43	0
035	Sodium (%)	0098	0.3600	0.0200	0.3662	0.0197	0.0098	85	-0.31	035.43	0
035	Sodium (%)	0682	0.3600	0.0000	0.3662	0.0197	0.0098	85	-0.31	035.43	0
035	Sodium (%)	0242	0.3600	0.0000	0.3662	0.0197	0.0098	85	-0.31	035.99	0
035	Sodium (%)	0948	0.3605	0.0030	0.3662	0.0197	0.0098	85	-0.29	035.31	0
035	Sodium (%)	2012	0.3605	0.0010	0.3662	0.0197	0.0098	85	-0.29	035.41	0
035	Sodium (%)	0263	0.3625	0.0090	0.3662	0.0197	0.0098	85	-0.19	035.42	0
035	Sodium (%)	0006	0.3625	0.0010	0.3662	0.0197	0.0098	85	-0.19	035.43	0
035	Sodium (%)	0407	0.3631	0.0061	0.3662	0.0197	0.0098	85	-0.16	035.53	0
035	Sodium (%)	0042	0.3635	0.0210	0.3662	0.0197	0.0098	85	-0.13	035.42	0
035	Sodium (%)	0019	0.3650	0.0100	0.3662	0.0197	0.0098	85	-0.06	035.41	0
035	Sodium (%)	0692	0.3650	0.0300	0.3662	0.0197	0.0098	85	-0.06	035.42	0
035	Sodium (%)	0870	0.3671	0.0099	0.3662	0.0197	0.0098	85	0.05	035.42	0
035	Sodium (%)	0598	0.3681	0.0016	0.3662	0.0197	0.0098	85	0.10	035.41	0
035	Sodium (%)	0358	0.3700	0.0200	0.3662	0.0197	0.0098	85	0.19	035.41	0
035	Sodium (%)	0190	0.3700	0.0000	0.3662	0.0197	0.0098	85	0.19	035.42	0
035	Sodium (%)	0265	0.3700	0.0000	0.3662	0.0197	0.0098	85	0.19	035.42	0
035	Sodium (%)	0693	0.3700	0.0000	0.3662	0.0197	0.0098	85	0.19	035.42	0
035	Sodium (%)	0297	0.3700	0.0200	0.3662	0.0197	0.0098	85	0.19	035.43	0
035	Sodium (%)	0300	0.3700	0.0200	0.3662	0.0197	0.0098	85	0.19	035.43	0
035	Sodium (%)	0610	0.3700	0.0000	0.3662	0.0197	0.0098	85	0.19	035.43	0
035	Sodium (%)	0083	0.3730	0.0040	0.3662	0.0197	0.0098	85	0.35	035.43	0
035	Sodium (%)	0202	0.3730	0.0100	0.3662	0.0197	0.0098	85	0.35	035.43	0
035	Sodium (%)	0876	0.3745	0.0110	0.3662	0.0197	0.0098	85	0.42	035.43	0
035	Sodium (%)	2246	0.3750	0.0100	0.3662	0.0197	0.0098	85	0.45	035.31	0
035	Sodium (%)	0098	0.3750	0.0100	0.3662	0.0197	0.0098	85	0.45	035.41	0
035	Sodium (%)	0294	0.3750	0.0100	0.3662	0.0197	0.0098	85	0.45	035.42	0
035	Sodium (%)	0968	0.3750	0.0020	0.3662	0.0197	0.0098	85	0.45	035.43	0
035	Sodium (%)	0123	0.3750	0.0100	0.3662	0.0197	0.0098	85	0.45	035.99	0
035	Sodium (%)	0017	0.3757	0.0035	0.3662	0.0197	0.0098	85	0.48	035.43	0
035	Sodium (%)	0142	0.3757	0.0240	0.3662	0.0197	0.0098	85	0.48	035.31	0
035	Sodium (%)	0148	0.3775	0.0090	0.3662	0.0197	0.0098	85	0.57	035.41	0
035	Sodium (%)	0035	0.3800	0.0000	0.3662	0.0197	0.0098	85	0.70	035.42	0
035	Sodium (%)	0504	0.3800	0.0200	0.3662	0.0197	0.0098	85	0.70	035.42	0
035	Sodium (%)	0726	0.3800	0.0000	0.3662	0.0197	0.0098	85	0.70	035.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
035	Sodium (%)	2375	0.3800	0.0000	0.3662	0.0197	0.0098	85	0.70	035.42	0
035	Sodium (%)	0035	0.3800	0.0000	0.3662	0.0197	0.0098	85	0.70	035.43	0
035	Sodium (%)	0353	0.3801	0.0151	0.3662	0.0197	0.0098	85	0.70	035.43	0
035	Sodium (%)	2314	0.3844	0.0030	0.3662	0.0197	0.0098	85	0.92	035.43	0
035	Sodium (%)	0038	0.3865	0.0250	0.3662	0.0197	0.0098	85	1.03	035.31	0
035	Sodium (%)	0529	0.3865	0.0050	0.3662	0.0197	0.0098	85	1.03	035.31	0
035	Sodium (%)	0186	0.3871	0.0079	0.3662	0.0197	0.0098	85	1.06	035.42	0
035	Sodium (%)	0870	0.3888	0.0036	0.3662	0.0197	0.0098	85	1.15	035.43	0
035	Sodium (%)	0226	0.3900	0.0000	0.3662	0.0197	0.0098	85	1.21	035.41	0
035	Sodium (%)	0366	0.3900	0.0600	0.3662	0.0197	0.0098	85	1.21	035.42	0
035	Sodium (%)	2089	0.3900	0.0000	0.3662	0.0197	0.0098	85	1.21	035.43	0
035	Sodium (%)	0510	0.3920	0.0040	0.3662	0.0197	0.0098	85	1.31	035.43	0
035	Sodium (%)	0723	0.3946	0.0284	0.3662	0.0197	0.0098	85	1.44	035.41	0
035	Sodium (%)	0675	0.3950	0.0100	0.3662	0.0197	0.0098	85	1.46	035.43	0
035	Sodium (%)	0354	0.4075	0.0010	0.3662	0.0197	0.0098	85	2.09	035.41	0
035	Sodium (%)	0278	0.4100	0.0000	0.3662	0.0197	0.0098	85	2.22	035.42	0
035	Sodium (%)	0186	0.4152	0.0062	0.3662	0.0197	0.0098	85	2.48	035.52	0
035	Sodium (%)	0626	0.4630	0.0060	0.3662	0.0197	0.0098	85	4.90	035.43	0
035	Sodium (%)	2188	0.4900	0.0000	0.3662	0.0197	0.0098	85	6.27	035.01	0
035	Sodium (%)	0045	0.3220	0.0900	0.3662	0.0197	0.0098	85	-2.24	035.42	1
035	Sodium (%)	0108	0.8250	0.0700	0.3662	0.0197	0.0098	85	23.23	035.05	2
036	Sulfur (%)	0171	0.2700	0.0200	0.3216	0.0230	0.0114	46	-2.24	036.42	0
036	Sulfur (%)	2129	0.2843	0.0021	0.3216	0.0230	0.0114	46	-1.62	036.42	0
036	Sulfur (%)	0242	0.2850	0.0100	0.3216	0.0230	0.0114	46	-1.59	036.99	0
036	Sulfur (%)	0294	0.2950	0.0100	0.3216	0.0230	0.0114	46	-1.16	036.42	0
036	Sulfur (%)	0366	0.2950	0.0100	0.3216	0.0230	0.0114	46	-1.16	036.42	0
036	Sulfur (%)	0357	0.2970	0.0013	0.3216	0.0230	0.0114	46	-1.07	036.42	0
036	Sulfur (%)	0870	0.2982	0.0203	0.3216	0.0230	0.0114	46	-1.02	036.42	0
036	Sulfur (%)	0265	0.3000	0.0200	0.3216	0.0230	0.0114	46	-0.94	036.42	0
036	Sulfur (%)	0297	0.3000	0.0200	0.3216	0.0230	0.0114	46	-0.94	036.43	0
036	Sulfur (%)	0045	0.3015	0.0190	0.3216	0.0230	0.0114	46	-0.87	036.42	0
036	Sulfur (%)	0407	0.3043	0.0015	0.3216	0.0230	0.0114	46	-0.75	036.43	0
036	Sulfur (%)	0035	0.3050	0.0100	0.3216	0.0230	0.0114	46	-0.72	036.42	0
036	Sulfur (%)	0164	0.3050	0.0100	0.3216	0.0230	0.0114	46	-0.72	036.42	0
036	Sulfur (%)	0083	0.3050	0.0100	0.3216	0.0230	0.0114	46	-0.72	036.43	0
036	Sulfur (%)	0001	0.3065	0.0270	0.3216	0.0230	0.0114	46	-0.66	036.04	0
036	Sulfur (%)	0693	0.3100	0.0200	0.3216	0.0230	0.0114	46	-0.50	036.42	0
036	Sulfur (%)	0042	0.3120	0.0140	0.3216	0.0230	0.0114	46	-0.42	036.42	0
036	Sulfur (%)	2326	0.3150	0.0300	0.3216	0.0230	0.0114	46	-0.29	036.43	0
036	Sulfur (%)	0553	0.3155	0.0010	0.3216	0.0230	0.0114	46	-0.27	036.53	0
036	Sulfur (%)	0345	0.3165	0.0010	0.3216	0.0230	0.0114	46	-0.22	036.43	0
036	Sulfur (%)	0708	0.3195	0.0010	0.3216	0.0230	0.0114	46	-0.09	036.42	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
036	Sulfur (%)	0726	0.3200	0.0000	0.3216	0.0230	0.0114	46	-0.07	036.42	0
036	Sulfur (%)	0629	0.3200	0.0200	0.3216	0.0230	0.0114	46	-0.07	036.43	0
036	Sulfur (%)	0186	0.3220	0.0020	0.3216	0.0230	0.0114	46	0.02	036.42	0
036	Sulfur (%)	0148	0.3225	0.0130	0.3216	0.0230	0.0114	46	0.04	036.04	0
036	Sulfur (%)	0229	0.3250	0.0100	0.3216	0.0230	0.0114	46	0.15	036.04	0
036	Sulfur (%)	2375	0.3250	0.0100	0.3216	0.0230	0.0114	46	0.15	036.42	0
036	Sulfur (%)	0948	0.3265	0.0012	0.3216	0.0230	0.0114	46	0.21	036.04	0
036	Sulfur (%)	0098	0.3282	0.0101	0.3216	0.0230	0.0114	46	0.28	036.43	0
036	Sulfur (%)	0190	0.3300	0.0000	0.3216	0.0230	0.0114	46	0.36	036.42	0
036	Sulfur (%)	0353	0.3350	0.0100	0.3216	0.0230	0.0114	46	0.58	036.43	0
036	Sulfur (%)	0407	0.3353	0.0286	0.3216	0.0230	0.0114	46	0.59	036.42	0
036	Sulfur (%)	0202	0.3365	0.0290	0.3216	0.0230	0.0114	46	0.65	036.43	0
036	Sulfur (%)	0510	0.3400	0.0200	0.3216	0.0230	0.0114	46	0.80	036.43	0
036	Sulfur (%)	0035	0.3400	0.0000	0.3216	0.0230	0.0114	46	0.80	036.43	0
036	Sulfur (%)	0123	0.3400	0.0000	0.3216	0.0230	0.0114	46	0.80	036.99	0
036	Sulfur (%)	0038	0.3420	0.0120	0.3216	0.0230	0.0114	46	0.88	036.00	0
036	Sulfur (%)	0870	0.3435	0.0119	0.3216	0.0230	0.0114	46	0.95	036.43	0
036	Sulfur (%)	0876	0.3439	0.0138	0.3216	0.0230	0.0114	46	0.97	036.43	0
036	Sulfur (%)	0226	0.3450	0.0100	0.3216	0.0230	0.0114	46	1.02	036.04	0
036	Sulfur (%)	2089	0.3450	0.0100	0.3216	0.0230	0.0114	46	1.02	036.43	0
036	Sulfur (%)	0278	0.3500	0.0200	0.3216	0.0230	0.0114	46	1.23	036.42	0
036	Sulfur (%)	0682	0.3500	0.0000	0.3216	0.0230	0.0114	46	1.23	036.43	0
036	Sulfur (%)	0169	0.3550	0.0100	0.3216	0.0230	0.0114	46	1.45	036.43	0
036	Sulfur (%)	0186	0.3568	0.0024	0.3216	0.0230	0.0114	46	1.53	036.52	0
036	Sulfur (%)	0066	0.3600	0.0200	0.3216	0.0230	0.0114	46	1.67	036.43	0
036	Sulfur (%)	0692	0.3000	0.0600	0.3216	0.0230	0.0114	46	-0.94	036.42	1
037	Zinc (ppm)	2326	118.0	13.79	157.1	12.03	4.556	87	-3.25	037.43	0
037	Zinc (ppm)	0066	119.9	3.889	157.1	12.03	4.556	87	-3.09	037.44	0
037	Zinc (ppm)	0723	120.6	2.596	157.1	12.03	4.556	87	-3.03	037.41	0
037	Zinc (ppm)	0123	131.0	0.0000	157.1	12.03	4.556	87	-2.17	037.99	0
037	Zinc (ppm)	0529	134.1	0.1000	157.1	12.03	4.556	87	-1.91	037.31	0
037	Zinc (ppm)	0511	137.5	7.000	157.1	12.03	4.556	87	-1.63	037.43	0
037	Zinc (ppm)	0870	140.0	3.200	157.1	12.03	4.556	87	-1.42	037.43	0
037	Zinc (ppm)	0242	140.5	1.000	157.1	12.03	4.556	87	-1.38	037.99	0
037	Zinc (ppm)	0870	141.9	2.000	157.1	12.03	4.556	87	-1.26	037.42	0
037	Zinc (ppm)	0520	142.2	7.239	157.1	12.03	4.556	87	-1.23	037.41	0
037	Zinc (ppm)	0848	143.0	7.730	157.1	12.03	4.556	87	-1.17	037.41	0
037	Zinc (ppm)	0358	143.2	10.04	157.1	12.03	4.556	87	-1.15	037.41	0
037	Zinc (ppm)	0045	144.0	8.000	157.1	12.03	4.556	87	-1.09	037.42	0
037	Zinc (ppm)	0651	145.4	5.238	157.1	12.03	4.556	87	-0.97	037.31	0
037	Zinc (ppm)	0354	145.4	0.7500	157.1	12.03	4.556	87	-0.97	037.41	0
037	Zinc (ppm)	0968	145.5	1.410	157.1	12.03	4.556	87	-0.97	037.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
037	Zinc (ppm)	0263	145.8	3.236	157.1	12.03	4.556	87	-0.93	037.42	0
037	Zinc (ppm)	0619	146.0	4.000	157.1	12.03	4.556	87	-0.92	037.41	0
037	Zinc (ppm)	0265	147.5	9.000	157.1	12.03	4.556	87	-0.80	037.42	0
037	Zinc (ppm)	0164	148.0	0.0000	157.1	12.03	4.556	87	-0.75	037.44	0
037	Zinc (ppm)	2129	149.3	5.300	157.1	12.03	4.556	87	-0.65	037.42	0
037	Zinc (ppm)	0510	149.5	1.000	157.1	12.03	4.556	87	-0.63	037.43	0
037	Zinc (ppm)	0019	150.4	7.500	157.1	12.03	4.556	87	-0.56	037.41	0
037	Zinc (ppm)	2192	150.5	5.000	157.1	12.03	4.556	87	-0.55	037.43	0
037	Zinc (ppm)	0049	150.6	16.57	157.1	12.03	4.556	87	-0.54	037.41	0
037	Zinc (ppm)	2012	151.2	0.6500	157.1	12.03	4.556	87	-0.49	037.41	0
037	Zinc (ppm)	0226	151.5	1.990	157.1	12.03	4.556	87	-0.46	037.43	0
037	Zinc (ppm)	2188	151.6	0.0000	157.1	12.03	4.556	87	-0.45	037.34	0
037	Zinc (ppm)	0629	152.0	4.000	157.1	12.03	4.556	87	-0.42	037.31	0
037	Zinc (ppm)	0598	152.6	0.3693	157.1	12.03	4.556	87	-0.37	037.41	0
037	Zinc (ppm)	0693	153.3	3.180	157.1	12.03	4.556	87	-0.31	037.42	0
037	Zinc (ppm)	0939	153.6	0.1600	157.1	12.03	4.556	87	-0.29	037.31	0
037	Zinc (ppm)	0098	154.0	8.100	157.1	12.03	4.556	87	-0.26	037.41	0
037	Zinc (ppm)	0910	154.0	2.000	157.1	12.03	4.556	87	-0.26	037.41	0
037	Zinc (ppm)	0692	154.0	4.000	157.1	12.03	4.556	87	-0.26	037.42	0
037	Zinc (ppm)	0848	154.1	1.150	157.1	12.03	4.556	87	-0.25	037.43	0
037	Zinc (ppm)	0512	155.6	8.700	157.1	12.03	4.556	87	-0.13	037.43	0
037	Zinc (ppm)	0675	156.3	3.020	157.1	12.03	4.556	87	-0.06	037.43	0
037	Zinc (ppm)	0229	156.5	8.000	157.1	12.03	4.556	87	-0.05	037.41	0
037	Zinc (ppm)	0047	156.6	0.5600	157.1	12.03	4.556	87	-0.04	037.52	0
037	Zinc (ppm)	2196	157.0	0.0000	157.1	12.03	4.556	87	-0.01	037.31	0
037	Zinc (ppm)	2319	157.0	4.000	157.1	12.03	4.556	87	-0.01	037.42	0
037	Zinc (ppm)	0169	157.0	4.000	157.1	12.03	4.556	87	-0.01	037.43	0
037	Zinc (ppm)	0407	157.2	6.969	157.1	12.03	4.556	87	0.01	037.43	0
037	Zinc (ppm)	0948	157.3	3.820	157.1	12.03	4.556	87	0.02	037.33	0
037	Zinc (ppm)	0366	157.5	17.00	157.1	12.03	4.556	87	0.03	037.42	0
037	Zinc (ppm)	0202	158.5	5.000	157.1	12.03	4.556	87	0.12	037.43	0
037	Zinc (ppm)	0407	158.9	5.786	157.1	12.03	4.556	87	0.15	037.53	0
037	Zinc (ppm)	0629	159.0	6.000	157.1	12.03	4.556	87	0.16	037.43	0
037	Zinc (ppm)	0294	159.1	0.6200	157.1	12.03	4.556	87	0.17	037.42	0
037	Zinc (ppm)	0357	159.1	0.0090	157.1	12.03	4.556	87	0.17	037.42	0
037	Zinc (ppm)	0674	159.2	0.0000	157.1	12.03	4.556	87	0.17	037.41	0
037	Zinc (ppm)	0682	159.2	0.0000	157.1	12.03	4.556	87	0.17	037.43	0
037	Zinc (ppm)	2022	159.4	2.450	157.1	12.03	4.556	87	0.19	037.33	0
037	Zinc (ppm)	0083	159.9	2.680	157.1	12.03	4.556	87	0.23	037.43	0
037	Zinc (ppm)	0345	159.9	1.200	157.1	12.03	4.556	87	0.23	037.43	0
037	Zinc (ppm)	2089	160.0	9.110	157.1	12.03	4.556	87	0.24	037.43	0
037	Zinc (ppm)	0010	160.1	15.80	157.1	12.03	4.556	87	0.25	037.33	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
037	Zinc (ppm)	0001	161.0	1.100	157.1	12.03	4.556	87	0.32	037.31	0
037	Zinc (ppm)	0004	161.2	10.56	157.1	12.03	4.556	87	0.34	037.41	0
037	Zinc (ppm)	0171	161.6	2.700	157.1	12.03	4.556	87	0.37	037.41	0
037	Zinc (ppm)	0226	161.8	1.280	157.1	12.03	4.556	87	0.39	037.41	0
037	Zinc (ppm)	0009	163.5	3.000	157.1	12.03	4.556	87	0.53	037.42	0
037	Zinc (ppm)	0148	163.8	2.817	157.1	12.03	4.556	87	0.56	037.41	0
037	Zinc (ppm)	0504	164.5	7.000	157.1	12.03	4.556	87	0.62	037.42	0
037	Zinc (ppm)	0027	164.6	12.23	157.1	12.03	4.556	87	0.62	037.43	0
037	Zinc (ppm)	0017	164.6	1.453	157.1	12.03	4.556	87	0.63	037.43	0
037	Zinc (ppm)	0407	166.1	5.798	157.1	12.03	4.556	87	0.75	037.41	0
037	Zinc (ppm)	0190	167.1	1.200	157.1	12.03	4.556	87	0.83	037.42	0
037	Zinc (ppm)	0553	168.0	10.00	157.1	12.03	4.556	87	0.91	037.53	0
037	Zinc (ppm)	2246	168.4	1.290	157.1	12.03	4.556	87	0.94	037.31	0
037	Zinc (ppm)	0505	168.9	7.000	157.1	12.03	4.556	87	0.98	037.43	0
037	Zinc (ppm)	0035	170.0	2.000	157.1	12.03	4.556	87	1.07	037.42	0
037	Zinc (ppm)	0042	170.0	6.000	157.1	12.03	4.556	87	1.07	037.42	0
037	Zinc (ppm)	0100	170.0	20.00	157.1	12.03	4.556	87	1.07	037.99	0
037	Zinc (ppm)	0626	171.0	0.7700	157.1	12.03	4.556	87	1.16	037.43	0
037	Zinc (ppm)	0035	171.5	3.000	157.1	12.03	4.556	87	1.20	037.43	0
037	Zinc (ppm)	0353	172.5	5.000	157.1	12.03	4.556	87	1.28	037.43	0
037	Zinc (ppm)	0610	173.0	8.000	157.1	12.03	4.556	87	1.32	037.43	0
037	Zinc (ppm)	0098	173.1	1.300	157.1	12.03	4.556	87	1.33	037.43	0
037	Zinc (ppm)	0656	174.2	0.0000	157.1	12.03	4.556	87	1.42	037.31	0
037	Zinc (ppm)	0186	175.5	1.000	157.1	12.03	4.556	87	1.53	037.52	0
037	Zinc (ppm)	0038	177.5	7.000	157.1	12.03	4.556	87	1.70	037.32	0
037	Zinc (ppm)	0876	177.8	4.500	157.1	12.03	4.556	87	1.72	037.43	0
037	Zinc (ppm)	0726	179.3	0.9000	157.1	12.03	4.556	87	1.84	037.42	0
037	Zinc (ppm)	0278	180.8	6.600	157.1	12.03	4.556	87	1.97	037.42	0
037	Zinc (ppm)	2375	188.5	7.000	157.1	12.03	4.556	87	2.61	037.42	0
037	Zinc (ppm)	0297	146.5	25.00	157.1	12.03	4.556	87	-0.88	037.43	1
037	Zinc (ppm)	0033	151.5	27.00	157.1	12.03	4.556	87	-0.46	037.43	1
038	Molybdenum (ppm)	0345	1.635	0.0500	2.042	0.2085	0.1367	21	-1.95	038.43	0
038	Molybdenum (ppm)	0083	1.685	0.0100	2.042	0.2085	0.1367	21	-1.71	038.43	0
038	Molybdenum (ppm)	0366	1.725	0.4500	2.042	0.2085	0.1367	21	-1.52	038.42	0
038	Molybdenum (ppm)	0171	1.825	0.0100	2.042	0.2085	0.1367	21	-1.04	038.41	0
038	Molybdenum (ppm)	0226	1.960	0.0000	2.042	0.2085	0.1367	21	-0.39	038.41	0
038	Molybdenum (ppm)	0407	1.968	0.1033	2.042	0.2085	0.1367	21	-0.35	038.41	0
038	Molybdenum (ppm)	0186	1.980	0.0800	2.042	0.2085	0.1367	21	-0.30	038.52	0
038	Molybdenum (ppm)	0297	2.000	0.4000	2.042	0.2085	0.1367	21	-0.20	038.43	0
038	Molybdenum (ppm)	0510	2.000	0.0000	2.042	0.2085	0.1367	21	-0.20	038.43	0
038	Molybdenum (ppm)	0202	2.005	0.0900	2.042	0.2085	0.1367	21	-0.18	038.43	0
038	Molybdenum (ppm)	0407	2.075	0.0191	2.042	0.2085	0.1367	21	0.16	038.43	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
038	Molybdenum (ppm)	0164	2.075	0.0220	2.042	0.2085	0.1367	21	0.16	038.53	0
038	Molybdenum (ppm)	0169	2.080	0.0200	2.042	0.2085	0.1367	21	0.18	038.43	0
038	Molybdenum (ppm)	0407	2.096	0.0857	2.042	0.2085	0.1367	21	0.26	038.53	0
038	Molybdenum (ppm)	0553	2.125	0.0700	2.042	0.2085	0.1367	21	0.40	038.53	0
038	Molybdenum (ppm)	0045	2.135	0.3700	2.042	0.2085	0.1367	21	0.45	038.42	0
038	Molybdenum (ppm)	0265	2.185	0.4700	2.042	0.2085	0.1367	21	0.69	038.42	0
038	Molybdenum (ppm)	0629	2.205	0.0100	2.042	0.2085	0.1367	21	0.78	038.43	0
038	Molybdenum (ppm)	0148	2.265	0.2100	2.042	0.2085	0.1367	21	1.07	038.41	0
038	Molybdenum (ppm)	0278	2.450	0.4000	2.042	0.2085	0.1367	21	1.96	038.42	0
038	Molybdenum (ppm)	0123	2.610	0.0000	2.042	0.2085	0.1367	21	2.73	038.99	0
040	Barium (ppm)	0407	9.895	0.1866				1		040.53	0
041	Vanadium (ppm)	0553	0.5025	0.0110				2		041.53	0
041	Vanadium (ppm)	0407	0.5866	0.0084				2		041.53	0
042	Chloride (%)	0619	0.4280	0.0000	0.5074	0.0593	0.0113	8	-1.34	042.00	0
042	Chloride (%)	0226	0.4800	0.0000	0.5074	0.0593	0.0113	8	-0.46	042.00	0
042	Chloride (%)	0948	0.4800	0.0000	0.5074	0.0593	0.0113	8	-0.46	042.00	0
042	Chloride (%)	0629	0.4900	0.0000	0.5074	0.0593	0.0113	8	-0.29	042.00	0
042	Chloride (%)	2326	0.5000	0.0000	0.5074	0.0593	0.0113	8	-0.13	042.00	0
042	Chloride (%)	0297	0.5250	0.0300	0.5074	0.0593	0.0113	8	0.30	042.99	0
042	Chloride (%)	2199	0.5600	0.0200	0.5074	0.0593	0.0113	8	0.89	042.00	0
042	Chloride (%)	2200	0.6300	0.0400	0.5074	0.0593	0.0113	8	2.07	042.00	0
099	Menadione (form) (ppm)	0619	1.145	0.0900				1		099.01	0
101	Choline Chloride (ppm)	0160	1,470	260.0				2		101.99	0
101	Choline Chloride (ppm)	0227	1,720	40.00				2		101.99	0
102	Niacin (ppm)	2319	62.00	3.400				2		102.02	0
102	Niacin (ppm)	0227	96.55	0.1000				2		102.01	0
103	Pantothenic Acid (ppm)	0010	39.15	1.900				2		103.02	0
103	Pantothenic Acid (ppm)	0227	39.95	1.300				2		103.01	0
104	Riboflavin (ppm)	2319	11.65	0.3000	13.77	1.518	0.2725	4		104.03	0
104	Riboflavin (ppm)	0227	13.70	0.2000	13.77	1.518	0.2725	4		104.03	0
104	Riboflavin (ppm)	0939	14.84	0.3900	13.77	1.518	0.2725	4		104.03	0
104	Riboflavin (ppm)	0010	14.90	0.2000	13.77	1.518	0.2725	4		104.03	0
105	Thiamine (ppm)	2319	2.400	0.0000				3		105.00	0
105	Thiamine (ppm)	0010	3.400	0.4000				3		105.00	0
105	Thiamine (ppm)	0227	4.330	0.0200				3		105.00	0
106	Vitamin A (KU / kg)	0676	8.470	0.0800	13.43	4.522	1.810	11	-1.10	106.02	0
106	Vitamin A (KU / kg)	0910	9.100	2.600	13.43	4.522	1.810	11	-0.96	106.02	0
106	Vitamin A (KU / kg)	2319	9.660	0.5200	13.43	4.522	1.810	11	-0.83	106.02	0
106	Vitamin A (KU / kg)	0610	11.35	0.3000	13.43	4.522	1.810	11	-0.46	106.02	0
106	Vitamin A (KU / kg)	0227	12.75	0.5000	13.43	4.522	1.810	11	-0.15	106.00	0
106	Vitamin A (KU / kg)	0004	12.76	2.190	13.43	4.522	1.810	11	-0.15	106.02	0
106	Vitamin A (KU / kg)	0148	13.46	0.3650	13.43	4.522	1.810	11	0.01	106.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
106	Vitamin A (KU / kg)	0010	14.85	1.610	13.43	4.522	1.810	11	0.31	106.02	0
106	Vitamin A (KU / kg)	0512	14.92	0.8500	13.43	4.522	1.810	11	0.33	106.02	0
106	Vitamin A (KU / kg)	0870	20.53	0.0000	13.43	4.522	1.810	11	1.57	106.02	0
106	Vitamin A (KU / kg)	0098	22.85	10.90	13.43	4.522	1.810	11	2.08	106.01	0
106	Vitamin A (KU / kg)	0638	32.45	25.10	13.43	4.522	1.810	11	4.21	106.02	1
107	Vitamin B12 (ppb)	0227	9.150	0.0600				1		107.00	0
108	Vitamin D3 (KU / kg)	2319	2.980	0.6400	3.228	0.2988	0.3100	4		108.02	0
108	Vitamin D3 (KU / kg)	0610	3.000	0.0000	3.228	0.2988	0.3100	4		108.02	0
108	Vitamin D3 (KU / kg)	0010	3.320	0.2000	3.228	0.2988	0.3100	4		108.99	0
108	Vitamin D3 (KU / kg)	0227	3.610	0.4000	3.228	0.2988	0.3100	4		108.02	0
108	Vitamin D3 (KU / kg)	0870	5.415	3.638	3.228	0.2988	0.3100	4		108.02	2
109	Vitamin E (IU / kg)	0638	87.00	124.0	112.9	19.13	17.52	11	-1.35	109.02	0
109	Vitamin E (IU / kg)	0098	94.55	28.90	112.9	19.13	17.52	11	-0.96	109.02	0
109	Vitamin E (IU / kg)	2193	99.48	2.210	112.9	19.13	17.52	11	-0.70	109.02	0
109	Vitamin E (IU / kg)	0910	100.3	7.400	112.9	19.13	17.52	11	-0.66	109.02	0
109	Vitamin E (IU / kg)	0148	104.7	0.0000	112.9	19.13	17.52	11	-0.43	109.02	0
109	Vitamin E (IU / kg)	0610	109.0	8.000	112.9	19.13	17.52	11	-0.20	109.02	0
109	Vitamin E (IU / kg)	2188	123.6	0.0000	112.9	19.13	17.52	11	0.56	109.02	0
109	Vitamin E (IU / kg)	2319	127.1	1.100	112.9	19.13	17.52	11	0.74	109.02	0
109	Vitamin E (IU / kg)	0227	130.5	1.000	112.9	19.13	17.52	11	0.92	109.02	0
109	Vitamin E (IU / kg)	0676	132.2	13.30	112.9	19.13	17.52	11	1.01	109.02	0
109	Vitamin E (IU / kg)	0870	133.4	6.821	112.9	19.13	17.52	11	1.07	109.02	0
109	Vitamin E (IU / kg)	0171	3,080	140.0	112.9	19.13	17.52	11	155.13	109.99	2
111	Vitamin C, Phosphorylated (ppm)	0227	< 11.4					0		111.00	5
112	Pyridoxine (µg / g)	0010	1.150	0.1000				3		112.01	0
112	Pyridoxine (µg / g)	2319	3.755	0.1500				3		112.01	0
112	Pyridoxine (µg / g)	0227	4.335	0.2100				3		112.01	0
113	Folic Acid (ppm)	2319	1.210	0.0400				2		113.02	0
113	Folic Acid (ppm)	0227	1.470	0.0200				2		113.01	0
114	Biotin (ppm)	0227	0.5805	0.0210				1		114.01	0
115	Non Protein N (NPN) (%)	0027	0.9855	0.1730	1.402	0.5271	0.0441	8	-0.79	115.99	0
115	Non Protein N (NPN) (%)	0505	1.030	0.0400	1.402	0.5271	0.0441	8	-0.71	115.99	0
115	Non Protein N (NPN) (%)	0001	1.100	0.0000	1.402	0.5271	0.0441	8	-0.57	115.00	0
115	Non Protein N (NPN) (%)	0171	1.145	0.0100	1.402	0.5271	0.0441	8	-0.49	115.00	0
115	Non Protein N (NPN) (%)	2192	1.280	0.0000	1.402	0.5271	0.0441	8	-0.23	115.01	0
115	Non Protein N (NPN) (%)	2129	1.449	0.0398	1.402	0.5271	0.0441	8	0.09	115.99	0
115	Non Protein N (NPN) (%)	0407	2.035	0.0100	1.402	0.5271	0.0441	8	1.20	115.00	0
115	Non Protein N (NPN) (%)	0098	7.240	0.0800	1.402	0.5271	0.0441	8	11.08	115.00	0
118	Peroxide value (meq/kg)	2081	5.430	1.320				1		118.99	0
120	Alanine (%)	0171	0.7100	0.0200	0.8244	0.0563	0.0091	28	-2.03	120.00	0
120	Alanine (%)	0407	0.7400	0.0240	0.8244	0.0563	0.0091	28	-1.50	120.05	0
120	Alanine (%)	2196	0.7450	0.0000	0.8244	0.0563	0.0091	28	-1.41	120.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
120	Alanine (%)	2319	0.7450	0.0100	0.8244	0.0563	0.0091	28	-1.41	120.05	0
120	Alanine (%)	0676	0.7500	0.0000	0.8244	0.0563	0.0091	28	-1.32	120.05	0
120	Alanine (%)	0353	0.7650	0.0100	0.8244	0.0563	0.0091	28	-1.06	120.00	0
120	Alanine (%)	0675	0.8050	0.0100	0.8244	0.0563	0.0091	28	-0.34	120.00	0
120	Alanine (%)	0148	0.8055	0.0110	0.8244	0.0563	0.0091	28	-0.34	120.05	0
120	Alanine (%)	0968	0.8100	0.0000	0.8244	0.0563	0.0091	28	-0.26	120.00	0
120	Alanine (%)	0504	0.8200	0.0000	0.8244	0.0563	0.0091	28	-0.08	120.00	0
120	Alanine (%)	0610	0.8200	0.0000	0.8244	0.0563	0.0091	28	-0.08	120.05	0
120	Alanine (%)	0226	0.8217	0.0098	0.8244	0.0563	0.0091	28	-0.05	120.00	0
120	Alanine (%)	0941	0.8230	0.0000	0.8244	0.0563	0.0091	28	-0.03	120.00	0
120	Alanine (%)	0354	0.8265	0.0050	0.8244	0.0563	0.0091	28	0.04	120.00	0
120	Alanine (%)	2059	0.8300	0.0020	0.8244	0.0563	0.0091	28	0.10	120.00	0
120	Alanine (%)	0571	0.8345	0.0090	0.8244	0.0563	0.0091	28	0.18	120.00	0
120	Alanine (%)	0682	0.8350	0.0000	0.8244	0.0563	0.0091	28	0.19	120.00	0
120	Alanine (%)	2188	0.8380	0.0000	0.8244	0.0563	0.0091	28	0.24	120.05	0
120	Alanine (%)	0644	0.8415	0.0050	0.8244	0.0563	0.0091	28	0.30	120.00	0
120	Alanine (%)	0910	0.8450	0.0300	0.8244	0.0563	0.0091	28	0.37	120.00	0
120	Alanine (%)	0859	0.8520	0.0129	0.8244	0.0563	0.0091	28	0.49	120.00	0
120	Alanine (%)	0227	0.8550	0.0100	0.8244	0.0563	0.0091	28	0.54	120.99	0
120	Alanine (%)	0870	0.8670	0.0040	0.8244	0.0563	0.0091	28	0.76	120.05	0
120	Alanine (%)	0684	0.8750	0.0300	0.8244	0.0563	0.0091	28	0.90	120.00	0
120	Alanine (%)	0652	0.8800	0.0200	0.8244	0.0563	0.0091	28	0.99	120.00	0
120	Alanine (%)	2246	0.9000	0.0200	0.8244	0.0563	0.0091	28	1.34	120.05	0
120	Alanine (%)	0939	0.9050	0.0100	0.8244	0.0563	0.0091	28	1.43	120.00	0
120	Alanine (%)	0626	0.9805	0.0010	0.8244	0.0563	0.0091	28	2.77	120.05	0
121	Arginine (%)	0682	1.137	0.0000	1.235	0.0462	0.0184	27	-2.13	121.00	0
121	Arginine (%)	0626	1.166	0.0190	1.235	0.0462	0.0184	27	-1.51	121.05	0
121	Arginine (%)	0148	1.170	0.0200	1.235	0.0462	0.0184	27	-1.41	121.05	0
121	Arginine (%)	0675	1.175	0.0100	1.235	0.0462	0.0184	27	-1.30	121.00	0
121	Arginine (%)	2188	1.183	0.0000	1.235	0.0462	0.0184	27	-1.13	121.05	0
121	Arginine (%)	0407	1.188	0.0420	1.235	0.0462	0.0184	27	-1.02	121.05	0
121	Arginine (%)	0226	1.214	0.0176	1.235	0.0462	0.0184	27	-0.46	121.00	0
121	Arginine (%)	2319	1.215	0.0500	1.235	0.0462	0.0184	27	-0.44	121.05	0
121	Arginine (%)	0968	1.220	0.0200	1.235	0.0462	0.0184	27	-0.33	121.00	0
121	Arginine (%)	0227	1.220	0.0000	1.235	0.0462	0.0184	27	-0.33	121.99	0
121	Arginine (%)	2059	1.222	0.0020	1.235	0.0462	0.0184	27	-0.29	121.00	0
121	Arginine (%)	0353	1.230	0.0400	1.235	0.0462	0.0184	27	-0.11	121.00	0
121	Arginine (%)	2196	1.232	0.0000	1.235	0.0462	0.0184	27	-0.07	121.05	0
121	Arginine (%)	0504	1.235	0.0100	1.235	0.0462	0.0184	27	0.00	121.00	0
121	Arginine (%)	0676	1.245	0.0500	1.235	0.0462	0.0184	27	0.21	121.05	0
121	Arginine (%)	0939	1.250	0.0200	1.235	0.0462	0.0184	27	0.32	121.00	0
121	Arginine (%)	0941	1.256	0.0000	1.235	0.0462	0.0184	27	0.45	121.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
121	Arginine (%)	0354	1.260	0.0130	1.235	0.0462	0.0184	27	0.53	121.00	0
121	Arginine (%)	0571	1.267	0.0030	1.235	0.0462	0.0184	27	0.68	121.00	0
121	Arginine (%)	0870	1.270	0.0330	1.235	0.0462	0.0184	27	0.74	121.05	0
121	Arginine (%)	0910	1.270	0.0000	1.235	0.0462	0.0184	27	0.75	121.00	0
121	Arginine (%)	0610	1.270	0.0200	1.235	0.0462	0.0184	27	0.75	121.05	0
121	Arginine (%)	0644	1.273	0.0070	1.235	0.0462	0.0184	27	0.81	121.00	0
121	Arginine (%)	0684	1.280	0.0400	1.235	0.0462	0.0184	27	0.97	121.00	0
121	Arginine (%)	0859	1.280	0.0311	1.235	0.0462	0.0184	27	0.97	121.00	0
121	Arginine (%)	0652	1.290	0.0200	1.235	0.0462	0.0184	27	1.19	121.00	0
121	Arginine (%)	2246	1.325	0.0300	1.235	0.0462	0.0184	27	1.95	121.05	0
121	Arginine (%)	0171	1.155	0.1100	1.235	0.0462	0.0184	27	-1.74	121.00	1
122	Aspartic (%)	0407	1.240	0.0790	1.464	0.0587	0.0197	28	-3.83	122.05	0
122	Aspartic (%)	0682	1.285	0.0000	1.464	0.0587	0.0197	28	-3.05	122.00	0
122	Aspartic (%)	0676	1.285	0.0300	1.464	0.0587	0.0197	28	-3.05	122.05	0
122	Aspartic (%)	0675	1.380	0.0200	1.464	0.0587	0.0197	28	-1.43	122.00	0
122	Aspartic (%)	0626	1.420	0.0170	1.464	0.0587	0.0197	28	-0.76	122.05	0
122	Aspartic (%)	2188	1.430	0.0000	1.464	0.0587	0.0197	28	-0.58	122.05	0
122	Aspartic (%)	2319	1.430	0.0200	1.464	0.0587	0.0197	28	-0.58	122.05	0
122	Aspartic (%)	0859	1.431	0.0088	1.464	0.0587	0.0197	28	-0.56	122.00	0
122	Aspartic (%)	0968	1.435	0.0300	1.464	0.0587	0.0197	28	-0.50	122.00	0
122	Aspartic (%)	2059	1.445	0.0060	1.464	0.0587	0.0197	28	-0.33	122.00	0
122	Aspartic (%)	0226	1.446	0.0240	1.464	0.0587	0.0197	28	-0.31	122.00	0
122	Aspartic (%)	0941	1.455	0.0000	1.464	0.0587	0.0197	28	-0.16	122.00	0
122	Aspartic (%)	0354	1.459	0.0230	1.464	0.0587	0.0197	28	-0.10	122.00	0
122	Aspartic (%)	0504	1.460	0.0000	1.464	0.0587	0.0197	28	-0.07	122.00	0
122	Aspartic (%)	0353	1.475	0.0100	1.464	0.0587	0.0197	28	0.18	122.00	0
122	Aspartic (%)	0684	1.475	0.0300	1.464	0.0587	0.0197	28	0.18	122.00	0
122	Aspartic (%)	0910	1.475	0.0100	1.464	0.0587	0.0197	28	0.18	122.00	0
122	Aspartic (%)	0571	1.479	0.0060	1.464	0.0587	0.0197	28	0.25	122.00	0
122	Aspartic (%)	0644	1.480	0.0070	1.464	0.0587	0.0197	28	0.26	122.00	0
122	Aspartic (%)	0610	1.490	0.0200	1.464	0.0587	0.0197	28	0.44	122.05	0
122	Aspartic (%)	2196	1.496	0.0000	1.464	0.0587	0.0197	28	0.54	122.05	0
122	Aspartic (%)	2246	1.505	0.0500	1.464	0.0587	0.0197	28	0.69	122.05	0
122	Aspartic (%)	0939	1.510	0.0000	1.464	0.0587	0.0197	28	0.78	122.00	0
122	Aspartic (%)	0227	1.520	0.0200	1.464	0.0587	0.0197	28	0.95	122.99	0
122	Aspartic (%)	0148	1.530	0.0000	1.464	0.0587	0.0197	28	1.12	122.05	0
122	Aspartic (%)	0652	1.540	0.0200	1.464	0.0587	0.0197	28	1.29	122.00	0
122	Aspartic (%)	0870	1.575	0.0500	1.464	0.0587	0.0197	28	1.89	122.05	0
122	Aspartic (%)	0171	1.585	0.0700	1.464	0.0587	0.0197	28	2.06	122.00	0
124	Cysteine/Cystine (%)	2246	0.2350	0.0100	0.3046	0.0247	0.0048	28	-2.81	124.05	0
124	Cysteine/Cystine (%)	2188	0.2540	0.0000	0.3046	0.0247	0.0048	28	-2.04	124.05	0
124	Cysteine/Cystine (%)	0227	0.2700	0.0200	0.3046	0.0247	0.0048	28	-1.40	124.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
124	Cysteine/Cystine (%)	0859	0.2730	0.0024	0.3046	0.0247	0.0048	28	-1.28	124.00	0
124	Cysteine/Cystine (%)	2193	0.2790	0.0040	0.3046	0.0247	0.0048	28	-1.03	124.05	0
124	Cysteine/Cystine (%)	0652	0.2900	0.0000	0.3046	0.0247	0.0048	28	-0.59	124.00	0
124	Cysteine/Cystine (%)	2319	0.2900	0.0000	0.3046	0.0247	0.0048	28	-0.59	124.05	0
124	Cysteine/Cystine (%)	0941	0.2920	0.0000	0.3046	0.0247	0.0048	28	-0.51	124.00	0
124	Cysteine/Cystine (%)	0353	0.2950	0.0100	0.3046	0.0247	0.0048	28	-0.39	124.00	0
124	Cysteine/Cystine (%)	0968	0.3000	0.0000	0.3046	0.0247	0.0048	28	-0.19	124.00	0
124	Cysteine/Cystine (%)	0676	0.3000	0.0000	0.3046	0.0247	0.0048	28	-0.19	124.05	0
124	Cysteine/Cystine (%)	2059	0.3005	0.0010	0.3046	0.0247	0.0048	28	-0.17	124.00	0
124	Cysteine/Cystine (%)	0644	0.3025	0.0010	0.3046	0.0247	0.0048	28	-0.08	124.00	0
124	Cysteine/Cystine (%)	0870	0.3035	0.0110	0.3046	0.0247	0.0048	28	-0.04	124.05	0
124	Cysteine/Cystine (%)	2196	0.3050	0.0000	0.3046	0.0247	0.0048	28	0.02	124.05	0
124	Cysteine/Cystine (%)	0226	0.3080	0.0013	0.3046	0.0247	0.0048	28	0.14	124.00	0
124	Cysteine/Cystine (%)	0571	0.3080	0.0060	0.3046	0.0247	0.0048	28	0.14	124.00	0
124	Cysteine/Cystine (%)	0910	0.3100	0.0000	0.3046	0.0247	0.0048	28	0.22	124.00	0
124	Cysteine/Cystine (%)	0939	0.3100	0.0000	0.3046	0.0247	0.0048	28	0.22	124.00	0
124	Cysteine/Cystine (%)	0354	0.3115	0.0050	0.3046	0.0247	0.0048	28	0.28	124.00	0
124	Cysteine/Cystine (%)	0148	0.3120	0.0020	0.3046	0.0247	0.0048	28	0.30	124.05	0
124	Cysteine/Cystine (%)	0504	0.3150	0.0100	0.3046	0.0247	0.0048	28	0.42	124.00	0
124	Cysteine/Cystine (%)	0171	0.3300	0.0200	0.3046	0.0247	0.0048	28	1.03	124.00	0
124	Cysteine/Cystine (%)	0675	0.3300	0.0200	0.3046	0.0247	0.0048	28	1.03	124.00	0
124	Cysteine/Cystine (%)	0684	0.3350	0.0100	0.3046	0.0247	0.0048	28	1.23	124.00	0
124	Cysteine/Cystine (%)	0610	0.3400	0.0000	0.3046	0.0247	0.0048	28	1.43	124.05	0
124	Cysteine/Cystine (%)	2246	0.3800	0.0000	0.3046	0.0247	0.0048	28	3.05	124.99	0
124	Cysteine/Cystine (%)	0682	0.4160	0.0000	0.3046	0.0247	0.0048	28	4.50	124.00	0
124	Cysteine/Cystine (%)	0407	0.2915	0.0315	0.3046	0.0247	0.0048	28	-0.53	124.05	1
125	Glutamic (%)	0407	2.679	0.0340	3.100	0.1308	0.0414	28	-3.22	125.05	0
125	Glutamic (%)	0675	2.840	0.0600	3.100	0.1308	0.0414	28	-1.98	125.00	0
125	Glutamic (%)	0676	2.900	0.1400	3.100	0.1308	0.0414	28	-1.53	125.05	0
125	Glutamic (%)	0682	2.955	0.0000	3.100	0.1308	0.0414	28	-1.11	125.00	0
125	Glutamic (%)	2319	2.955	0.0500	3.100	0.1308	0.0414	28	-1.11	125.05	0
125	Glutamic (%)	0968	3.005	0.0500	3.100	0.1308	0.0414	28	-0.72	125.00	0
125	Glutamic (%)	0859	3.022	0.0224	3.100	0.1308	0.0414	28	-0.59	125.00	0
125	Glutamic (%)	0626	3.024	0.0670	3.100	0.1308	0.0414	28	-0.58	125.05	0
125	Glutamic (%)	2059	3.054	0.0060	3.100	0.1308	0.0414	28	-0.35	125.00	0
125	Glutamic (%)	2188	3.069	0.0000	3.100	0.1308	0.0414	28	-0.23	125.05	0
125	Glutamic (%)	0353	3.075	0.0700	3.100	0.1308	0.0414	28	-0.19	125.00	0
125	Glutamic (%)	0504	3.075	0.0100	3.100	0.1308	0.0414	28	-0.19	125.00	0
125	Glutamic (%)	0941	3.100	0.0000	3.100	0.1308	0.0414	28	0.00	125.00	0
125	Glutamic (%)	2196	3.105	0.0000	3.100	0.1308	0.0414	28	0.04	125.05	0
125	Glutamic (%)	0571	3.110	0.0350	3.100	0.1308	0.0414	28	0.08	125.00	0
125	Glutamic (%)	2246	3.115	0.0900	3.100	0.1308	0.0414	28	0.12	125.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
125	Glutamic (%)	0171	3.125	0.0700	3.100	0.1308	0.0414	28	0.19	125.00	0
125	Glutamic (%)	0652	3.125	0.0300	3.100	0.1308	0.0414	28	0.19	125.00	0
125	Glutamic (%)	0227	3.135	0.0100	3.100	0.1308	0.0414	28	0.27	125.99	0
125	Glutamic (%)	0684	3.140	0.0800	3.100	0.1308	0.0414	28	0.31	125.00	0
125	Glutamic (%)	0910	3.165	0.0700	3.100	0.1308	0.0414	28	0.50	125.00	0
125	Glutamic (%)	0644	3.170	0.0160	3.100	0.1308	0.0414	28	0.54	125.00	0
125	Glutamic (%)	0354	3.174	0.0020	3.100	0.1308	0.0414	28	0.57	125.00	0
125	Glutamic (%)	0610	3.255	0.0500	3.100	0.1308	0.0414	28	1.19	125.05	0
125	Glutamic (%)	0939	3.265	0.0300	3.100	0.1308	0.0414	28	1.26	125.00	0
125	Glutamic (%)	0226	3.270	0.0050	3.100	0.1308	0.0414	28	1.30	125.00	0
125	Glutamic (%)	0148	3.375	0.0700	3.100	0.1308	0.0414	28	2.11	125.05	0
125	Glutamic (%)	0870	3.411	0.0920	3.100	0.1308	0.0414	28	2.38	125.05	0
126	Glycine (%)	0682	0.6650	0.0000	0.7490	0.0268	0.0086	27	-3.14	126.00	0
126	Glycine (%)	0148	0.7180	0.0180	0.7490	0.0268	0.0086	27	-1.16	126.05	0
126	Glycine (%)	2196	0.7190	0.0000	0.7490	0.0268	0.0086	27	-1.12	126.05	0
126	Glycine (%)	0353	0.7200	0.0000	0.7490	0.0268	0.0086	27	-1.08	126.00	0
126	Glycine (%)	0675	0.7200	0.0200	0.7490	0.0268	0.0086	27	-1.08	126.00	0
126	Glycine (%)	2319	0.7250	0.0100	0.7490	0.0268	0.0086	27	-0.90	126.05	0
126	Glycine (%)	0171	0.7300	0.0200	0.7490	0.0268	0.0086	27	-0.71	126.00	0
126	Glycine (%)	0676	0.7300	0.0400	0.7490	0.0268	0.0086	27	-0.71	126.05	0
126	Glycine (%)	0968	0.7350	0.0100	0.7490	0.0268	0.0086	27	-0.52	126.00	0
126	Glycine (%)	0859	0.7370	0.0108	0.7490	0.0268	0.0086	27	-0.45	126.00	0
126	Glycine (%)	2059	0.7405	0.0010	0.7490	0.0268	0.0086	27	-0.32	126.00	0
126	Glycine (%)	0684	0.7450	0.0100	0.7490	0.0268	0.0086	27	-0.15	126.00	0
126	Glycine (%)	0226	0.7472	0.0005	0.7490	0.0268	0.0086	27	-0.07	126.00	0
126	Glycine (%)	0504	0.7500	0.0000	0.7490	0.0268	0.0086	27	0.04	126.00	0
126	Glycine (%)	0354	0.7510	0.0080	0.7490	0.0268	0.0086	27	0.07	126.00	0
126	Glycine (%)	0941	0.7520	0.0000	0.7490	0.0268	0.0086	27	0.11	126.00	0
126	Glycine (%)	0644	0.7595	0.0010	0.7490	0.0268	0.0086	27	0.39	126.00	0
126	Glycine (%)	0652	0.7600	0.0000	0.7490	0.0268	0.0086	27	0.41	126.00	0
126	Glycine (%)	0610	0.7600	0.0000	0.7490	0.0268	0.0086	27	0.41	126.05	0
126	Glycine (%)	0571	0.7615	0.0070	0.7490	0.0268	0.0086	27	0.46	126.00	0
126	Glycine (%)	0910	0.7650	0.0100	0.7490	0.0268	0.0086	27	0.60	126.00	0
126	Glycine (%)	0227	0.7650	0.0100	0.7490	0.0268	0.0086	27	0.60	126.99	0
126	Glycine (%)	0407	0.7770	0.0340	0.7490	0.0268	0.0086	27	1.04	126.05	0
126	Glycine (%)	2246	0.7800	0.0200	0.7490	0.0268	0.0086	27	1.15	126.05	0
126	Glycine (%)	0939	0.8000	0.0000	0.7490	0.0268	0.0086	27	1.90	126.00	0
126	Glycine (%)	0870	0.8145	0.0030	0.7490	0.0268	0.0086	27	2.44	126.05	0
126	Glycine (%)	2188	0.9010	0.0000	0.7490	0.0268	0.0086	27	5.67	126.05	0
126	Glycine (%)	0626	1.495	0.0340	0.7490	0.0268	0.0086	27	27.83	126.05	2
127	Histidine (%)	2188	0.3000	0.0000	0.4334	0.0364	0.0068	28	-3.66	127.05	0
127	Histidine (%)	0676	0.3250	0.0100	0.4334	0.0364	0.0068	28	-2.97	127.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
127	Histidine (%)	0148	0.3770	0.0180	0.4334	0.0364	0.0068	28	-1.55	127.05	0
127	Histidine (%)	0682	0.3830	0.0000	0.4334	0.0364	0.0068	28	-1.38	127.00	0
127	Histidine (%)	0859	0.4117	0.0086	0.4334	0.0364	0.0068	28	-0.60	127.00	0
127	Histidine (%)	2059	0.4165	0.0010	0.4334	0.0364	0.0068	28	-0.46	127.00	0
127	Histidine (%)	0407	0.4190	0.0080	0.4334	0.0364	0.0068	28	-0.39	127.05	0
127	Histidine (%)	0171	0.4200	0.0200	0.4334	0.0364	0.0068	28	-0.37	127.00	0
127	Histidine (%)	0968	0.4200	0.0000	0.4334	0.0364	0.0068	28	-0.37	127.00	0
127	Histidine (%)	0353	0.4250	0.0100	0.4334	0.0364	0.0068	28	-0.23	127.00	0
127	Histidine (%)	0354	0.4250	0.0040	0.4334	0.0364	0.0068	28	-0.23	127.00	0
127	Histidine (%)	0675	0.4250	0.0100	0.4334	0.0364	0.0068	28	-0.23	127.00	0
127	Histidine (%)	0610	0.4250	0.0100	0.4334	0.0364	0.0068	28	-0.23	127.05	0
127	Histidine (%)	2319	0.4250	0.0100	0.4334	0.0364	0.0068	28	-0.23	127.05	0
127	Histidine (%)	0571	0.4340	0.0000	0.4334	0.0364	0.0068	28	0.02	127.00	0
127	Histidine (%)	0910	0.4350	0.0100	0.4334	0.0364	0.0068	28	0.04	127.00	0
127	Histidine (%)	0227	0.4350	0.0100	0.4334	0.0364	0.0068	28	0.04	127.99	0
127	Histidine (%)	0941	0.4390	0.0000	0.4334	0.0364	0.0068	28	0.15	127.00	0
127	Histidine (%)	0504	0.4400	0.0000	0.4334	0.0364	0.0068	28	0.18	127.00	0
127	Histidine (%)	0939	0.4400	0.0000	0.4334	0.0364	0.0068	28	0.18	127.00	0
127	Histidine (%)	0644	0.4425	0.0010	0.4334	0.0364	0.0068	28	0.25	127.00	0
127	Histidine (%)	0684	0.4500	0.0200	0.4334	0.0364	0.0068	28	0.46	127.00	0
127	Histidine (%)	0870	0.4640	0.0060	0.4334	0.0364	0.0068	28	0.84	127.05	0
127	Histidine (%)	0652	0.4800	0.0000	0.4334	0.0364	0.0068	28	1.28	127.00	0
127	Histidine (%)	0626	0.4810	0.0160	0.4334	0.0364	0.0068	28	1.31	127.05	0
127	Histidine (%)	2196	0.4870	0.0000	0.4334	0.0364	0.0068	28	1.47	127.05	0
127	Histidine (%)	2246	0.5050	0.0100	0.4334	0.0364	0.0068	28	1.96	127.05	0
127	Histidine (%)	0226	0.5089	0.0077	0.4334	0.0364	0.0068	28	2.07	127.00	0
128	Isoleucine (%)	0682	0.4610	0.0000	0.5836	0.0409	0.0112	27	-3.00	128.00	0
128	Isoleucine (%)	2319	0.4800	0.0400	0.5836	0.0409	0.0112	27	-2.54	128.05	0
128	Isoleucine (%)	0171	0.5050	0.0300	0.5836	0.0409	0.0112	27	-1.92	128.00	0
128	Isoleucine (%)	0676	0.5250	0.0100	0.5836	0.0409	0.0112	27	-1.43	128.05	0
128	Isoleucine (%)	0353	0.5500	0.0000	0.5836	0.0409	0.0112	27	-0.82	128.00	0
128	Isoleucine (%)	0684	0.5500	0.0200	0.5836	0.0409	0.0112	27	-0.82	128.00	0
128	Isoleucine (%)	0968	0.5550	0.0100	0.5836	0.0409	0.0112	27	-0.70	128.00	0
128	Isoleucine (%)	0939	0.5700	0.0000	0.5836	0.0409	0.0112	27	-0.33	128.00	0
128	Isoleucine (%)	0354	0.5730	0.0040	0.5836	0.0409	0.0112	27	-0.26	128.00	0
128	Isoleucine (%)	0675	0.5750	0.0100	0.5836	0.0409	0.0112	27	-0.21	128.00	0
128	Isoleucine (%)	0407	0.5750	0.0340	0.5836	0.0409	0.0112	27	-0.21	128.05	0
128	Isoleucine (%)	0226	0.5807	0.0081	0.5836	0.0409	0.0112	27	-0.07	128.00	0
128	Isoleucine (%)	0652	0.5850	0.0300	0.5836	0.0409	0.0112	27	0.03	128.00	0
128	Isoleucine (%)	0910	0.5900	0.0000	0.5836	0.0409	0.0112	27	0.16	128.00	0
128	Isoleucine (%)	2059	0.5905	0.0030	0.5836	0.0409	0.0112	27	0.17	128.00	0
128	Isoleucine (%)	0941	0.5920	0.0000	0.5836	0.0409	0.0112	27	0.20	128.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
128	Isoleucine (%)	0644	0.5935	0.0050	0.5836	0.0409	0.0112	27	0.24	128.00	0
128	Isoleucine (%)	2188	0.5970	0.0000	0.5836	0.0409	0.0112	27	0.33	128.05	0
128	Isoleucine (%)	0227	0.6000	0.0000	0.5836	0.0409	0.0112	27	0.40	128.99	0
128	Isoleucine (%)	0571	0.6055	0.0030	0.5836	0.0409	0.0112	27	0.54	128.00	0
128	Isoleucine (%)	0610	0.6100	0.0200	0.5836	0.0409	0.0112	27	0.65	128.05	0
128	Isoleucine (%)	2246	0.6100	0.0000	0.5836	0.0409	0.0112	27	0.65	128.05	0
128	Isoleucine (%)	0870	0.6130	0.0300	0.5836	0.0409	0.0112	27	0.72	128.05	0
128	Isoleucine (%)	0626	0.6210	0.0120	0.5836	0.0409	0.0112	27	0.91	128.05	0
128	Isoleucine (%)	2196	0.6400	0.0000	0.5836	0.0409	0.0112	27	1.38	128.05	0
128	Isoleucine (%)	0859	0.6449	0.0126	0.5836	0.0409	0.0112	27	1.50	128.00	0
128	Isoleucine (%)	0504	0.6500	0.0200	0.5836	0.0409	0.0112	27	1.62	128.00	0
128	Isoleucine (%)	0148	0.5490	0.0700	0.5836	0.0409	0.0112	27	-0.85	128.05	1
129	Leucine (%)	2319	1.045	0.0300	1.207	0.0529	0.0213	28	-3.06	129.05	0
129	Leucine (%)	2188	1.123	0.0000	1.207	0.0529	0.0213	28	-1.59	129.05	0
129	Leucine (%)	0682	1.129	0.0000	1.207	0.0529	0.0213	28	-1.47	129.00	0
129	Leucine (%)	0171	1.135	0.0900	1.207	0.0529	0.0213	28	-1.36	129.00	0
129	Leucine (%)	0676	1.135	0.0300	1.207	0.0529	0.0213	28	-1.36	129.05	0
129	Leucine (%)	0407	1.152	0.0590	1.207	0.0529	0.0213	28	-1.05	129.05	0
129	Leucine (%)	0675	1.170	0.0200	1.207	0.0529	0.0213	28	-0.70	129.00	0
129	Leucine (%)	0968	1.185	0.0100	1.207	0.0529	0.0213	28	-0.42	129.00	0
129	Leucine (%)	0148	1.185	0.0500	1.207	0.0529	0.0213	28	-0.42	129.05	0
129	Leucine (%)	0353	1.190	0.0400	1.207	0.0529	0.0213	28	-0.32	129.00	0
129	Leucine (%)	0684	1.195	0.0100	1.207	0.0529	0.0213	28	-0.23	129.00	0
129	Leucine (%)	0354	1.214	0.0210	1.207	0.0529	0.0213	28	0.12	129.00	0
129	Leucine (%)	0941	1.217	0.0000	1.207	0.0529	0.0213	28	0.19	129.00	0
129	Leucine (%)	2059	1.218	0.0050	1.207	0.0529	0.0213	28	0.20	129.00	0
129	Leucine (%)	2196	1.218	0.0000	1.207	0.0529	0.0213	28	0.21	129.05	0
129	Leucine (%)	0652	1.220	0.0200	1.207	0.0529	0.0213	28	0.25	129.00	0
129	Leucine (%)	0910	1.225	0.0100	1.207	0.0529	0.0213	28	0.34	129.00	0
129	Leucine (%)	0226	1.232	0.0135	1.207	0.0529	0.0213	28	0.48	129.00	0
129	Leucine (%)	0939	1.235	0.0100	1.207	0.0529	0.0213	28	0.53	129.00	0
129	Leucine (%)	0571	1.237	0.0210	1.207	0.0529	0.0213	28	0.56	129.00	0
129	Leucine (%)	0610	1.240	0.0200	1.207	0.0529	0.0213	28	0.62	129.05	0
129	Leucine (%)	0859	1.249	0.0179	1.207	0.0529	0.0213	28	0.79	129.00	0
129	Leucine (%)	0227	1.250	0.0200	1.207	0.0529	0.0213	28	0.81	129.99	0
129	Leucine (%)	0644	1.252	0.0070	1.207	0.0529	0.0213	28	0.84	129.00	0
129	Leucine (%)	2246	1.255	0.0100	1.207	0.0529	0.0213	28	0.91	129.05	0
129	Leucine (%)	0626	1.259	0.0210	1.207	0.0529	0.0213	28	0.97	129.05	0
129	Leucine (%)	0504	1.260	0.0000	1.207	0.0529	0.0213	28	1.00	129.00	0
129	Leucine (%)	0870	1.306	0.0620	1.207	0.0529	0.0213	28	1.87	129.05	0
130	L-Lysine (%)	0407	0.6030	0.0420	0.7147	0.0286	0.0118	28	-3.91	130.05	0
130	L-Lysine (%)	2188	0.6420	0.0000	0.7147	0.0286	0.0118	28	-2.54	130.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
130	L-Lysine (%)	0676	0.6850	0.0100	0.7147	0.0286	0.0118	28	-1.04	130.05	0
130	L-Lysine (%)	0684	0.6950	0.0100	0.7147	0.0286	0.0118	28	-0.69	130.00	0
130	L-Lysine (%)	0910	0.6950	0.0100	0.7147	0.0286	0.0118	28	-0.69	130.00	0
130	L-Lysine (%)	0968	0.6950	0.0100	0.7147	0.0286	0.0118	28	-0.69	130.00	0
130	L-Lysine (%)	0171	0.6950	0.0300	0.7147	0.0286	0.0118	28	-0.69	130.00	0
130	L-Lysine (%)	0941	0.6970	0.0000	0.7147	0.0286	0.0118	28	-0.62	130.00	0
130	L-Lysine (%)	0226	0.6984	0.0168	0.7147	0.0286	0.0118	28	-0.57	130.00	0
130	L-Lysine (%)	0682	0.7000	0.0000	0.7147	0.0286	0.0118	28	-0.51	130.00	0
130	L-Lysine (%)	0859	0.7001	0.0115	0.7147	0.0286	0.0118	28	-0.51	130.00	0
130	L-Lysine (%)	2059	0.7050	0.0020	0.7147	0.0286	0.0118	28	-0.34	130.00	0
130	L-Lysine (%)	2246	0.7050	0.0300	0.7147	0.0286	0.0118	28	-0.34	130.05	0
130	L-Lysine (%)	0354	0.7110	0.0120	0.7147	0.0286	0.0118	28	-0.13	130.00	0
130	L-Lysine (%)	0571	0.7145	0.0110	0.7147	0.0286	0.0118	28	-0.01	130.00	0
130	L-Lysine (%)	0675	0.7150	0.0100	0.7147	0.0286	0.0118	28	0.01	130.00	0
130	L-Lysine (%)	0939	0.7150	0.0100	0.7147	0.0286	0.0118	28	0.01	130.00	0
130	L-Lysine (%)	0610	0.7150	0.0100	0.7147	0.0286	0.0118	28	0.01	130.05	0
130	L-Lysine (%)	0626	0.7240	0.0100	0.7147	0.0286	0.0118	28	0.33	130.05	0
130	L-Lysine (%)	0644	0.7245	0.0010	0.7147	0.0286	0.0118	28	0.34	130.00	0
130	L-Lysine (%)	0353	0.7250	0.0100	0.7147	0.0286	0.0118	28	0.36	130.00	0
130	L-Lysine (%)	0652	0.7300	0.0200	0.7147	0.0286	0.0118	28	0.54	130.00	0
130	L-Lysine (%)	2196	0.7450	0.0000	0.7147	0.0286	0.0118	28	1.06	130.05	0
130	L-Lysine (%)	0504	0.7500	0.0000	0.7147	0.0286	0.0118	28	1.24	130.00	0
130	L-Lysine (%)	2246	0.7550	0.0300	0.7147	0.0286	0.0118	28	1.41	130.99	0
130	L-Lysine (%)	0870	0.7720	0.0040	0.7147	0.0286	0.0118	28	2.01	130.05	0
130	L-Lysine (%)	0227	0.7900	0.0200	0.7147	0.0286	0.0118	28	2.64	130.99	0
130	L-Lysine (%)	2319	0.8050	0.0100	0.7147	0.0286	0.0118	28	3.16	130.05	0
130	L-Lysine (%)	0148	0.6440	0.0800	0.7147	0.0286	0.0118	28	-2.47	130.05	1
131	Methionine (%)	2246	0.1850	0.0100	0.2508	0.0210	0.0081	29	-3.13	131.05	0
131	Methionine (%)	2246	0.1850	0.0300	0.2508	0.0210	0.0081	29	-3.13	131.99	0
131	Methionine (%)	0148	0.2190	0.0440	0.2508	0.0210	0.0081	29	-1.51	131.05	0
131	Methionine (%)	0226	0.2225	0.0030	0.2508	0.0210	0.0081	29	-1.35	131.00	0
131	Methionine (%)	2188	0.2310	0.0000	0.2508	0.0210	0.0081	29	-0.94	131.05	0
131	Methionine (%)	0968	0.2350	0.0100	0.2508	0.0210	0.0081	29	-0.75	131.00	0
131	Methionine (%)	0939	0.2400	0.0000	0.2508	0.0210	0.0081	29	-0.51	131.00	0
131	Methionine (%)	0626	0.2400	0.0000	0.2508	0.0210	0.0081	29	-0.51	131.05	0
131	Methionine (%)	2319	0.2400	0.0000	0.2508	0.0210	0.0081	29	-0.51	131.05	0
131	Methionine (%)	0682	0.2430	0.0000	0.2508	0.0210	0.0081	29	-0.37	131.00	0
131	Methionine (%)	0941	0.2440	0.0000	0.2508	0.0210	0.0081	29	-0.32	131.00	0
131	Methionine (%)	0610	0.2450	0.0100	0.2508	0.0210	0.0081	29	-0.28	131.05	0
131	Methionine (%)	2193	0.2475	0.0090	0.2508	0.0210	0.0081	29	-0.16	131.05	0
131	Methionine (%)	0354	0.2485	0.0090	0.2508	0.0210	0.0081	29	-0.11	131.00	0
131	Methionine (%)	2059	0.2525	0.0010	0.2508	0.0210	0.0081	29	0.08	131.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
131	Methionine (%)	0870	0.2525	0.0010	0.2508	0.0210	0.0081	29	0.08	131.05	0
131	Methionine (%)	0652	0.2550	0.0100	0.2508	0.0210	0.0081	29	0.20	131.00	0
131	Methionine (%)	0910	0.2550	0.0100	0.2508	0.0210	0.0081	29	0.20	131.00	0
131	Methionine (%)	0504	0.2600	0.0000	0.2508	0.0210	0.0081	29	0.44	131.00	0
131	Methionine (%)	0675	0.2600	0.0200	0.2508	0.0210	0.0081	29	0.44	131.00	0
131	Methionine (%)	0227	0.2600	0.0200	0.2508	0.0210	0.0081	29	0.44	131.99	0
131	Methionine (%)	0859	0.2601	0.0035	0.2508	0.0210	0.0081	29	0.44	131.00	0
131	Methionine (%)	0644	0.2635	0.0030	0.2508	0.0210	0.0081	29	0.61	131.00	0
131	Methionine (%)	0571	0.2700	0.0060	0.2508	0.0210	0.0081	29	0.92	131.00	0
131	Methionine (%)	0407	0.2713	0.0256	0.2508	0.0210	0.0081	29	0.98	131.05	0
131	Methionine (%)	2196	0.2760	0.0000	0.2508	0.0210	0.0081	29	1.20	131.05	0
131	Methionine (%)	0684	0.2800	0.0000	0.2508	0.0210	0.0081	29	1.39	131.00	0
131	Methionine (%)	0676	0.2800	0.0000	0.2508	0.0210	0.0081	29	1.39	131.05	0
131	Methionine (%)	0353	0.3050	0.0100	0.2508	0.0210	0.0081	29	2.58	131.00	0
131	Methionine (%)	0171	0.2900	0.0600	0.2508	0.0210	0.0081	29	1.87	131.00	1
132	Phenylalanine (%)	0682	0.6760	0.0000	0.7976	0.0349	0.0142	27	-3.49	132.00	0
132	Phenylalanine (%)	0148	0.6965	0.0310	0.7976	0.0349	0.0142	27	-2.90	132.05	0
132	Phenylalanine (%)	2188	0.7260	0.0000	0.7976	0.0349	0.0142	27	-2.05	132.05	0
132	Phenylalanine (%)	0171	0.7350	0.0300	0.7976	0.0349	0.0142	27	-1.80	132.00	0
132	Phenylalanine (%)	0968	0.7750	0.0100	0.7976	0.0349	0.0142	27	-0.65	132.00	0
132	Phenylalanine (%)	2059	0.7775	0.0030	0.7976	0.0349	0.0142	27	-0.58	132.00	0
132	Phenylalanine (%)	0652	0.7800	0.0000	0.7976	0.0349	0.0142	27	-0.50	132.00	0
132	Phenylalanine (%)	0939	0.7850	0.0100	0.7976	0.0349	0.0142	27	-0.36	132.00	0
132	Phenylalanine (%)	0610	0.7850	0.0100	0.7976	0.0349	0.0142	27	-0.36	132.05	0
132	Phenylalanine (%)	0676	0.7850	0.0300	0.7976	0.0349	0.0142	27	-0.36	132.05	0
132	Phenylalanine (%)	2196	0.7920	0.0000	0.7976	0.0349	0.0142	27	-0.16	132.05	0
132	Phenylalanine (%)	0941	0.7940	0.0000	0.7976	0.0349	0.0142	27	-0.10	132.00	0
132	Phenylalanine (%)	0675	0.7950	0.0300	0.7976	0.0349	0.0142	27	-0.07	132.00	0
132	Phenylalanine (%)	0227	0.7950	0.0100	0.7976	0.0349	0.0142	27	-0.07	132.99	0
132	Phenylalanine (%)	0407	0.8000	0.0100	0.7976	0.0349	0.0142	27	0.07	132.05	0
132	Phenylalanine (%)	2319	0.8000	0.0200	0.7976	0.0349	0.0142	27	0.07	132.05	0
132	Phenylalanine (%)	0571	0.8045	0.0050	0.7976	0.0349	0.0142	27	0.20	132.00	0
132	Phenylalanine (%)	0910	0.8050	0.0300	0.7976	0.0349	0.0142	27	0.21	132.00	0
132	Phenylalanine (%)	0644	0.8115	0.0010	0.7976	0.0349	0.0142	27	0.40	132.00	0
132	Phenylalanine (%)	0354	0.8175	0.0210	0.7976	0.0349	0.0142	27	0.57	132.00	0
132	Phenylalanine (%)	0859	0.8208	0.0176	0.7976	0.0349	0.0142	27	0.67	132.00	0
132	Phenylalanine (%)	0226	0.8233	0.0194	0.7976	0.0349	0.0142	27	0.74	132.00	0
132	Phenylalanine (%)	0504	0.8300	0.0000	0.7976	0.0349	0.0142	27	0.93	132.00	0
132	Phenylalanine (%)	0353	0.8350	0.0100	0.7976	0.0349	0.0142	27	1.07	132.00	0
132	Phenylalanine (%)	2246	0.8450	0.0300	0.7976	0.0349	0.0142	27	1.36	132.05	0
132	Phenylalanine (%)	0626	0.8485	0.0090	0.7976	0.0349	0.0142	27	1.46	132.05	0
132	Phenylalanine (%)	0870	0.8495	0.0470	0.7976	0.0349	0.0142	27	1.49	132.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
132	Phenylalanine (%)	0684	0.7500	0.0800	0.7976	0.0349	0.0142	27	-1.37	132.00	1
133	Proline (%)	0353	0.7950	0.0100	0.9202	0.0409	0.0202	28	-3.06	133.00	0
133	Proline (%)	0859	0.8045	0.0190	0.9202	0.0409	0.0202	28	-2.83	133.00	0
133	Proline (%)	0939	0.8150	0.0100	0.9202	0.0409	0.0202	28	-2.57	133.00	0
133	Proline (%)	0407	0.8780	0.0440	0.9202	0.0409	0.0202	28	-1.03	133.05	0
133	Proline (%)	0675	0.8800	0.0200	0.9202	0.0409	0.0202	28	-0.98	133.00	0
133	Proline (%)	0148	0.8880	0.0300	0.9202	0.0409	0.0202	28	-0.79	133.05	0
133	Proline (%)	0354	0.8885	0.0670	0.9202	0.0409	0.0202	28	-0.77	133.00	0
133	Proline (%)	2319	0.9050	0.0100	0.9202	0.0409	0.0202	28	-0.37	133.05	0
133	Proline (%)	2059	0.9075	0.0090	0.9202	0.0409	0.0202	28	-0.31	133.00	0
133	Proline (%)	0171	0.9100	0.0400	0.9202	0.0409	0.0202	28	-0.25	133.00	0
133	Proline (%)	0676	0.9100	0.0000	0.9202	0.0409	0.0202	28	-0.25	133.05	0
133	Proline (%)	0941	0.9110	0.0000	0.9202	0.0409	0.0202	28	-0.22	133.00	0
133	Proline (%)	0504	0.9150	0.0100	0.9202	0.0409	0.0202	28	-0.13	133.00	0
133	Proline (%)	2196	0.9170	0.0000	0.9202	0.0409	0.0202	28	-0.08	133.05	0
133	Proline (%)	0684	0.9200	0.0600	0.9202	0.0409	0.0202	28	0.00	133.00	0
133	Proline (%)	0610	0.9200	0.0000	0.9202	0.0409	0.0202	28	0.00	133.05	0
133	Proline (%)	0571	0.9205	0.0050	0.9202	0.0409	0.0202	28	0.01	133.00	0
133	Proline (%)	0968	0.9250	0.0100	0.9202	0.0409	0.0202	28	0.12	133.00	0
133	Proline (%)	2246	0.9250	0.0300	0.9202	0.0409	0.0202	28	0.12	133.05	0
133	Proline (%)	0644	0.9395	0.0050	0.9202	0.0409	0.0202	28	0.47	133.00	0
133	Proline (%)	2188	0.9530	0.0000	0.9202	0.0409	0.0202	28	0.80	133.05	0
133	Proline (%)	0626	0.9595	0.0030	0.9202	0.0409	0.0202	28	0.96	133.05	0
133	Proline (%)	0652	0.9650	0.0300	0.9202	0.0409	0.0202	28	1.09	133.00	0
133	Proline (%)	0227	0.9650	0.0900	0.9202	0.0409	0.0202	28	1.09	133.99	0
133	Proline (%)	0870	0.9665	0.0310	0.9202	0.0409	0.0202	28	1.13	133.05	0
133	Proline (%)	0910	0.9700	0.0200	0.9202	0.0409	0.0202	28	1.22	133.00	0
133	Proline (%)	0226	0.9743	0.0122	0.9202	0.0409	0.0202	28	1.32	133.00	0
133	Proline (%)	0682	0.9760	0.0000	0.9202	0.0409	0.0202	28	1.36	133.00	0
134	Serine (%)	0407	0.6060	0.0260	0.7583	0.0570	0.0103	27	-2.67	134.05	0
134	Serine (%)	0504	0.6450	0.0500	0.7583	0.0570	0.0103	27	-1.99	134.00	0
134	Serine (%)	2188	0.6530	0.0000	0.7583	0.0570	0.0103	27	-1.85	134.05	0
134	Serine (%)	0675	0.6850	0.0100	0.7583	0.0570	0.0103	27	-1.29	134.00	0
134	Serine (%)	0682	0.6940	0.0000	0.7583	0.0570	0.0103	27	-1.13	134.00	0
134	Serine (%)	0676	0.7100	0.0200	0.7583	0.0570	0.0103	27	-0.85	134.05	0
134	Serine (%)	2196	0.7290	0.0000	0.7583	0.0570	0.0103	27	-0.51	134.05	0
134	Serine (%)	0859	0.7326	0.0035	0.7583	0.0570	0.0103	27	-0.45	134.00	0
134	Serine (%)	0353	0.7450	0.0100	0.7583	0.0570	0.0103	27	-0.23	134.00	0
134	Serine (%)	0610	0.7450	0.0100	0.7583	0.0570	0.0103	27	-0.23	134.05	0
134	Serine (%)	0626	0.7470	0.0060	0.7583	0.0570	0.0103	27	-0.20	134.05	0
134	Serine (%)	0571	0.7545	0.0090	0.7583	0.0570	0.0103	27	-0.07	134.00	0
134	Serine (%)	2059	0.7550	0.0020	0.7583	0.0570	0.0103	27	-0.06	134.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
134	Serine (%)	0941	0.7570	0.0000	0.7583	0.0570	0.0103	27	-0.02	134.00	0
134	Serine (%)	0226	0.7622	0.0085	0.7583	0.0570	0.0103	27	0.07	134.00	0
134	Serine (%)	0968	0.7650	0.0100	0.7583	0.0570	0.0103	27	0.12	134.00	0
134	Serine (%)	0910	0.7750	0.0100	0.7583	0.0570	0.0103	27	0.29	134.00	0
134	Serine (%)	0171	0.7800	0.0200	0.7583	0.0570	0.0103	27	0.38	134.00	0
134	Serine (%)	0227	0.7950	0.0100	0.7583	0.0570	0.0103	27	0.64	134.99	0
134	Serine (%)	0354	0.7975	0.0010	0.7583	0.0570	0.0103	27	0.69	134.00	0
134	Serine (%)	0684	0.8000	0.0400	0.7583	0.0570	0.0103	27	0.73	134.00	0
134	Serine (%)	0644	0.8040	0.0020	0.7583	0.0570	0.0103	27	0.80	134.00	0
134	Serine (%)	0939	0.8100	0.0000	0.7583	0.0570	0.0103	27	0.91	134.00	0
134	Serine (%)	0652	0.8200	0.0200	0.7583	0.0570	0.0103	27	1.08	134.00	0
134	Serine (%)	0870	0.8275	0.0010	0.7583	0.0570	0.0103	27	1.22	134.05	0
134	Serine (%)	2319	0.8300	0.0000	0.7583	0.0570	0.0103	27	1.26	134.05	0
134	Serine (%)	2246	0.8350	0.0100	0.7583	0.0570	0.0103	27	1.35	134.05	0
134	Serine (%)	0148	0.8060	0.0840	0.7583	0.0570	0.0103	27	0.84	134.05	1
135	Threonine (%)	0407	0.4605	0.0230	0.5759	0.0318	0.0114	29	-3.63	135.05	0
135	Threonine (%)	0626	0.4950	0.0000	0.5759	0.0318	0.0114	29	-2.55	135.05	0
135	Threonine (%)	0682	0.5330	0.0000	0.5759	0.0318	0.0114	29	-1.35	135.00	0
135	Threonine (%)	0675	0.5350	0.0100	0.5759	0.0318	0.0114	29	-1.29	135.00	0
135	Threonine (%)	0676	0.5450	0.0100	0.5759	0.0318	0.0114	29	-0.97	135.05	0
135	Threonine (%)	2196	0.5460	0.0000	0.5759	0.0318	0.0114	29	-0.94	135.05	0
135	Threonine (%)	0226	0.5519	0.0072	0.5759	0.0318	0.0114	29	-0.76	135.00	0
135	Threonine (%)	2188	0.5540	0.0000	0.5759	0.0318	0.0114	29	-0.69	135.05	0
135	Threonine (%)	0353	0.5650	0.0100	0.5759	0.0318	0.0114	29	-0.34	135.00	0
135	Threonine (%)	0941	0.5690	0.0000	0.5759	0.0318	0.0114	29	-0.22	135.00	0
135	Threonine (%)	0171	0.5700	0.0400	0.5759	0.0318	0.0114	29	-0.19	135.00	0
135	Threonine (%)	0684	0.5700	0.0200	0.5759	0.0318	0.0114	29	-0.19	135.00	0
135	Threonine (%)	0504	0.5750	0.0100	0.5759	0.0318	0.0114	29	-0.03	135.00	0
135	Threonine (%)	0939	0.5750	0.0100	0.5759	0.0318	0.0114	29	-0.03	135.00	0
135	Threonine (%)	0610	0.5750	0.0100	0.5759	0.0318	0.0114	29	-0.03	135.05	0
135	Threonine (%)	2059	0.5755	0.0010	0.5759	0.0318	0.0114	29	-0.01	135.00	0
135	Threonine (%)	0148	0.5760	0.0500	0.5759	0.0318	0.0114	29	0.00	135.05	0
135	Threonine (%)	0859	0.5828	0.0038	0.5759	0.0318	0.0114	29	0.22	135.00	0
135	Threonine (%)	0571	0.5850	0.0000	0.5759	0.0318	0.0114	29	0.29	135.00	0
135	Threonine (%)	0910	0.5850	0.0100	0.5759	0.0318	0.0114	29	0.29	135.00	0
135	Threonine (%)	0968	0.5900	0.0200	0.5759	0.0318	0.0114	29	0.44	135.00	0
135	Threonine (%)	0644	0.5970	0.0040	0.5759	0.0318	0.0114	29	0.66	135.00	0
135	Threonine (%)	0652	0.6000	0.0000	0.5759	0.0318	0.0114	29	0.76	135.00	0
135	Threonine (%)	0227	0.6050	0.0100	0.5759	0.0318	0.0114	29	0.92	135.99	0
135	Threonine (%)	2319	0.6100	0.0200	0.5759	0.0318	0.0114	29	1.07	135.05	0
135	Threonine (%)	2246	0.6150	0.0300	0.5759	0.0318	0.0114	29	1.23	135.99	0
135	Threonine (%)	0354	0.6155	0.0010	0.5759	0.0318	0.0114	29	1.25	135.00	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
135	Threonine (%)	0870	0.6200	0.0200	0.5759	0.0318	0.0114	29	1.39	135.05	0
135	Threonine (%)	2246	0.6250	0.0100	0.5759	0.0318	0.0114	29	1.55	135.05	0
136	Tryptophan (%)	0504	0.1750	0.0100	0.2348	0.0348	0.0052	18	-1.72	136.99	0
136	Tryptophan (%)	0682	0.2020	0.0000	0.2348	0.0348	0.0052	18	-0.94	136.00	0
136	Tryptophan (%)	0644	0.2060	0.0040	0.2348	0.0348	0.0052	18	-0.83	136.01	0
136	Tryptophan (%)	0859	0.2083	0.0136	0.2348	0.0348	0.0052	18	-0.76	136.03	0
136	Tryptophan (%)	2196	0.2093	0.0000	0.2348	0.0348	0.0052	18	-0.73	136.03	0
136	Tryptophan (%)	2319	0.2100	0.0000	0.2348	0.0348	0.0052	18	-0.71	136.05	0
136	Tryptophan (%)	2059	0.2200	0.0040	0.2348	0.0348	0.0052	18	-0.42	136.03	0
136	Tryptophan (%)	0610	0.2250	0.0100	0.2348	0.0348	0.0052	18	-0.28	136.99	0
136	Tryptophan (%)	0571	0.2265	0.0030	0.2348	0.0348	0.0052	18	-0.24	136.03	0
136	Tryptophan (%)	0941	0.2310	0.0000	0.2348	0.0348	0.0052	18	-0.11	136.00	0
136	Tryptophan (%)	0939	0.2400	0.0000	0.2348	0.0348	0.0052	18	0.15	136.03	0
136	Tryptophan (%)	0226	0.2444	0.0048	0.2348	0.0348	0.0052	18	0.28	136.03	0
136	Tryptophan (%)	0353	0.2450	0.0100	0.2348	0.0348	0.0052	18	0.29	136.03	0
136	Tryptophan (%)	0227	0.2600	0.0000	0.2348	0.0348	0.0052	18	0.72	136.00	0
136	Tryptophan (%)	0407	0.2640	0.0020	0.2348	0.0348	0.0052	18	0.84	136.05	0
136	Tryptophan (%)	0870	0.2780	0.0120	0.2348	0.0348	0.0052	18	1.24	136.05	0
136	Tryptophan (%)	0684	0.3450	0.0100	0.2348	0.0348	0.0052	18	3.16	136.00	0
136	Tryptophan (%)	2246	0.3450	0.0100	0.2348	0.0348	0.0052	18	3.16	136.99	0
137	Tyrosine (%)	0682	0.3640	0.0000	0.5029	0.0649	0.0117	25	-2.14	137.00	0
137	Tyrosine (%)	0675	0.3950	0.0100	0.5029	0.0649	0.0117	25	-1.66	137.00	0
137	Tyrosine (%)	0910	0.4100	0.0200	0.5029	0.0649	0.0117	25	-1.43	137.00	0
137	Tyrosine (%)	0227	0.4400	0.0000	0.5029	0.0649	0.0117	25	-0.97	137.99	0
137	Tyrosine (%)	2246	0.4550	0.0300	0.5029	0.0649	0.0117	25	-0.74	137.05	0
137	Tyrosine (%)	0870	0.4550	0.0220	0.5029	0.0649	0.0117	25	-0.74	137.05	0
137	Tyrosine (%)	0407	0.4675	0.0010	0.5029	0.0649	0.0117	25	-0.55	137.05	0
137	Tyrosine (%)	0504	0.4850	0.0100	0.5029	0.0649	0.0117	25	-0.28	137.00	0
137	Tyrosine (%)	0354	0.4865	0.0310	0.5029	0.0649	0.0117	25	-0.25	137.00	0
137	Tyrosine (%)	0941	0.4870	0.0000	0.5029	0.0649	0.0117	25	-0.25	137.00	0
137	Tyrosine (%)	0684	0.4900	0.0200	0.5029	0.0649	0.0117	25	-0.20	137.00	0
137	Tyrosine (%)	0626	0.4915	0.0030	0.5029	0.0649	0.0117	25	-0.18	137.05	0
137	Tyrosine (%)	0968	0.4950	0.0100	0.5029	0.0649	0.0117	25	-0.12	137.00	0
137	Tyrosine (%)	0148	0.4985	0.0330	0.5029	0.0649	0.0117	25	-0.07	137.05	0
137	Tyrosine (%)	2319	0.5000	0.0000	0.5029	0.0649	0.0117	25	-0.05	137.05	0
137	Tyrosine (%)	0171	0.5100	0.0400	0.5029	0.0649	0.0117	25	0.11	137.00	0
137	Tyrosine (%)	2196	0.5350	0.0000	0.5029	0.0649	0.0117	25	0.49	137.05	0
137	Tyrosine (%)	0644	0.5375	0.0010	0.5029	0.0649	0.0117	25	0.53	137.00	0
137	Tyrosine (%)	0226	0.5486	0.0085	0.5029	0.0649	0.0117	25	0.70	137.00	0
137	Tyrosine (%)	0939	0.5550	0.0100	0.5029	0.0649	0.0117	25	0.80	137.00	0
137	Tyrosine (%)	0676	0.5650	0.0100	0.5029	0.0649	0.0117	25	0.96	137.05	0
137	Tyrosine (%)	2188	0.5720	0.0000	0.5029	0.0649	0.0117	25	1.07	137.05	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
137	Tyrosine (%)	2059	0.5875	0.0030	0.5029	0.0649	0.0117	25	1.30	137.00	0
137	Tyrosine (%)	0610	0.5900	0.0200	0.5029	0.0649	0.0117	25	1.34	137.05	0
137	Tyrosine (%)	0353	0.6050	0.0100	0.5029	0.0649	0.0117	25	1.57	137.00	0
138	Valine (%)	2319	0.6400	0.0600	0.7623	0.0381	0.0155	28	-3.21	138.05	0
138	Valine (%)	0171	0.6700	0.0200	0.7623	0.0381	0.0155	28	-2.42	138.00	0
138	Valine (%)	2188	0.7090	0.0000	0.7623	0.0381	0.0155	28	-1.40	138.05	0
138	Valine (%)	0682	0.7240	0.0000	0.7623	0.0381	0.0155	28	-1.01	138.00	0
138	Valine (%)	2196	0.7280	0.0000	0.7623	0.0381	0.0155	28	-0.90	138.05	0
138	Valine (%)	0684	0.7300	0.0400	0.7623	0.0381	0.0155	28	-0.85	138.00	0
138	Valine (%)	0676	0.7300	0.0000	0.7623	0.0381	0.0155	28	-0.85	138.05	0
138	Valine (%)	0354	0.7350	0.0000	0.7623	0.0381	0.0155	28	-0.72	138.00	0
138	Valine (%)	0941	0.7430	0.0000	0.7623	0.0381	0.0155	28	-0.51	138.00	0
138	Valine (%)	0407	0.7435	0.0410	0.7623	0.0381	0.0155	28	-0.49	138.05	0
138	Valine (%)	0968	0.7450	0.0100	0.7623	0.0381	0.0155	28	-0.45	138.00	0
138	Valine (%)	0353	0.7500	0.0400	0.7623	0.0381	0.0155	28	-0.32	138.00	0
138	Valine (%)	0910	0.7650	0.0100	0.7623	0.0381	0.0155	28	0.07	138.00	0
138	Valine (%)	2059	0.7650	0.0000	0.7623	0.0381	0.0155	28	0.07	138.00	0
138	Valine (%)	0644	0.7675	0.0050	0.7623	0.0381	0.0155	28	0.14	138.00	0
138	Valine (%)	0610	0.7700	0.0200	0.7623	0.0381	0.0155	28	0.20	138.05	0
138	Valine (%)	0148	0.7725	0.0470	0.7623	0.0381	0.0155	28	0.27	138.05	0
138	Valine (%)	0626	0.7740	0.0140	0.7623	0.0381	0.0155	28	0.31	138.05	0
138	Valine (%)	2246	0.7850	0.0100	0.7623	0.0381	0.0155	28	0.60	138.05	0
138	Valine (%)	0226	0.7856	0.0095	0.7623	0.0381	0.0155	28	0.61	138.00	0
138	Valine (%)	0652	0.7900	0.0200	0.7623	0.0381	0.0155	28	0.73	138.00	0
138	Valine (%)	0939	0.7900	0.0000	0.7623	0.0381	0.0155	28	0.73	138.00	0
138	Valine (%)	0675	0.7950	0.0100	0.7623	0.0381	0.0155	28	0.86	138.00	0
138	Valine (%)	0571	0.7985	0.0230	0.7623	0.0381	0.0155	28	0.95	138.00	0
138	Valine (%)	0859	0.8045	0.0146	0.7623	0.0381	0.0155	28	1.11	138.00	0
138	Valine (%)	0870	0.8050	0.0200	0.7623	0.0381	0.0155	28	1.12	138.05	0
138	Valine (%)	0504	0.8100	0.0200	0.7623	0.0381	0.0155	28	1.25	138.00	0
138	Valine (%)	0227	0.8300	0.0000	0.7623	0.0381	0.0155	28	1.78	138.99	0
139	Taurine (%)	0407	0.0100	0.0000	0.0960	0.0863	0.0000	4		139.05	0
139	Taurine (%)	0682	0.0540	0.0000	0.0960	0.0863	0.0000	4		139.00	0
139	Taurine (%)	0171	0.1100	0.0000	0.0960	0.0863	0.0000	4		139.00	0
139	Taurine (%)	0504	0.2100	0.0000	0.0960	0.0863	0.0000	4		139.00	0
139	Taurine (%)	0227	< 0.01		0.0960	0.0863	0.0000	4		139.99	5
160	Fructose (%)	0227	0.4050	0.0100				3		160.99	0
160	Fructose (%)	0297	0.4300	0.0200				3		160.10	0
160	Fructose (%)	0148	0.5250	0.0500				3		160.99	0
161	Galactose (%)	0297	0.0000	0.0000				0		161.10	4
162	Glucose (%)	0297	0.2850	0.0100				3		162.10	0
162	Glucose (%)	0148	0.3050	0.0900				3		162.99	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
162	Glucose (%)	0227	0.3350	0.0500				3		162.99	0
162	Glucose (%)	0689	5.350	0.1000				3		162.99	2
163	Lactose (%)	0939	1.830	0.0600				1		163.99	0
163	Lactose (%)	0297	0.0000	0.0000				1		163.10	4
163	Lactose (%)	0227	< 0.15					1		163.99	5
164	Maltose (%)	0148	0.1550	0.0300				2		164.99	0
164	Maltose (%)	0297	0.4250	0.0700				2		164.10	0
164	Maltose (%)	0227	< 0.15					2		164.99	5
165	Sucrose (%)	0297	2.340	0.1600				3		165.10	0
165	Sucrose (%)	0227	2.340	0.0200				3		165.99	0
165	Sucrose (%)	0148	2.515	0.0100				3		165.99	0
166	Raffinose (%)	0227	0.5750	0.0100				2		166.99	0
166	Raffinose (%)	0297	0.8250	0.1300				2		166.10	0
167	Stachyose (%)	0297	0.6900	0.0800				2		167.10	0
167	Stachyose (%)	0227	0.8300	0.0000				2		167.99	0
351	Chlortetracycline (ppm)	0148	2.520	0.4200	4.524	1.536	0.1550	4		351.05	0
351	Chlortetracycline (ppm)	0941	4.375	0.0000	4.524	1.536	0.1550	4		351.05	0
351	Chlortetracycline (ppm)	0910	5.000	0.0000	4.524	1.536	0.1550	4		351.03	0
351	Chlortetracycline (ppm)	0036	6.200	0.2000	4.524	1.536	0.1550	4		351.03	0
351	Chlortetracycline (ppm)	0218	5.349	1.046	4.524	1.536	0.1550	4		351.00	1
351	Chlortetracycline (ppm)	0038	< 5		4.524	1.536	0.1550	4		351.03	5
354	Decoquinatate (ppm)	0032	2.390	0.0200	27.52	2.428	0.9186	14	-10.35	354.02	0
354	Decoquinatate (ppm)	0047	24.09	0.3400	27.52	2.428	0.9186	14	-1.41	354.02	0
354	Decoquinatate (ppm)	0028	25.45	1.500	27.52	2.428	0.9186	14	-0.85	354.01	0
354	Decoquinatate (ppm)	0610	26.00	0.6000	27.52	2.428	0.9186	14	-0.63	354.04	0
354	Decoquinatate (ppm)	2113	26.50	1.000	27.52	2.428	0.9186	14	-0.42	354.04	0
354	Decoquinatate (ppm)	0910	27.00	0.2000	27.52	2.428	0.9186	14	-0.21	354.01	0
354	Decoquinatate (ppm)	0148	27.33	2.891	27.52	2.428	0.9186	14	-0.08	354.04	0
354	Decoquinatate (ppm)	0027	28.59	0.4100	27.52	2.428	0.9186	14	0.44	354.02	0
354	Decoquinatate (ppm)	0218	28.60	0.4000	27.52	2.428	0.9186	14	0.44	354.01	0
354	Decoquinatate (ppm)	0038	28.90	1.400	27.52	2.428	0.9186	14	0.57	354.01	0
354	Decoquinatate (ppm)	0001	29.15	0.1000	27.52	2.428	0.9186	14	0.67	354.01	0
354	Decoquinatate (ppm)	0010	29.30	0.2000	27.52	2.428	0.9186	14	0.73	354.01	0
354	Decoquinatate (ppm)	0036	29.35	0.3000	27.52	2.428	0.9186	14	0.75	354.02	0
354	Decoquinatate (ppm)	2192	51.85	3.500	27.52	2.428	0.9186	14	10.02	354.02	0
361	Lasalocid Sodium (ppm)	0910	< 0.1					0		361.02	5
365	Monensin (ppm)	0910	< 0.1					0		365.02	5
386	Tiamulin (ppm)	0910	1.350	0.3000				1		386.99	0
388	Tylosin (ppm)	0032	6.325	0.4500	23.80	13.22	2.567	5		388.03	0
388	Tylosin (ppm)	2192	19.00	0.4000	23.80	13.22	2.567	5		388.03	0
388	Tylosin (ppm)	0148	21.53	3.096	23.80	13.22	2.567	5		388.05	0
388	Tylosin (ppm)	0668	30.36	4.010	23.80	13.22	2.567	5		388.02	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
388	Tylosin (ppm)	0027	41.76	4.880	23.80	13.22	2.567	5		388.00	0
400	Water Activity (Units)	2174	0.5270	0.0060	0.5498	0.0158	0.0057	11	-1.44	400.01	0
400	Water Activity (Units)	2303	0.5380	0.0000	0.5498	0.0158	0.0057	11	-0.75	400.01	0
400	Water Activity (Units)	0942	0.5415	0.0070	0.5498	0.0158	0.0057	11	-0.53	400.01	0
400	Water Activity (Units)	2168	0.5425	0.0090	0.5498	0.0158	0.0057	11	-0.46	400.01	0
400	Water Activity (Units)	0589	0.5445	0.0070	0.5498	0.0158	0.0057	11	-0.34	400.01	0
400	Water Activity (Units)	0407	0.5481	0.0001	0.5498	0.0158	0.0057	11	-0.11	400.01	0
400	Water Activity (Units)	2268	0.5515	0.0150	0.5498	0.0158	0.0057	11	0.10	400.01	0
400	Water Activity (Units)	2073	0.5540	0.0080	0.5498	0.0158	0.0057	11	0.26	400.99	0
400	Water Activity (Units)	0123	0.5550	0.0100	0.5498	0.0158	0.0057	11	0.33	400.01	0
400	Water Activity (Units)	2109	0.5727	0.0001	0.5498	0.0158	0.0057	11	1.44	400.01	0
400	Water Activity (Units)	0885	0.6000	0.0000	0.5498	0.0158	0.0057	11	3.17	400.01	0
516	Arsenic, Total (As) (ppm)	0407	0.0306	0.0009	0.0530	0.0141	0.0013	5		516.53	0
516	Arsenic, Total (As) (ppm)	0553	0.0510	0.0004	0.0530	0.0141	0.0013	5		516.53	0
516	Arsenic, Total (As) (ppm)	0186	0.0570	0.0000	0.0530	0.0141	0.0013	5		516.52	0
516	Arsenic, Total (As) (ppm)	0227	0.0575	0.0030	0.0530	0.0141	0.0013	5		516.53	0
516	Arsenic, Total (As) (ppm)	2012	0.0690	0.0020	0.0530	0.0141	0.0013	5		516.53	0
516	Arsenic, Total (As) (ppm)	0045	0.0600	0.0200	0.0530	0.0141	0.0013	5		516.00	1
516	Arsenic, Total (As) (ppm)	0682	< 20		0.0530	0.0141	0.0013	5		516.43	5
518	Cadmium (ppm)	0171	0.0880	0.0020	0.1018	0.0113	0.0029	8	-1.22	518.41	0
518	Cadmium (ppm)	0407	0.0915	0.0050	0.1018	0.0113	0.0029	8	-0.91	518.43	0
518	Cadmium (ppm)	2012	0.0980	0.0020	0.1018	0.0113	0.0029	8	-0.34	518.53	0
518	Cadmium (ppm)	0407	0.0993	0.0002	0.1018	0.0113	0.0029	8	-0.22	518.41	0
518	Cadmium (ppm)	0407	0.0999	0.0048	0.1018	0.0113	0.0029	8	-0.17	518.53	0
518	Cadmium (ppm)	0553	0.1095	0.0030	0.1018	0.0113	0.0029	8	0.69	518.53	0
518	Cadmium (ppm)	0227	0.1115	0.0050	0.1018	0.0113	0.0029	8	0.86	518.53	0
518	Cadmium (ppm)	0186	0.1165	0.0010	0.1018	0.0113	0.0029	8	1.31	518.52	0
518	Cadmium (ppm)	0682	< 0.2		0.1018	0.0113	0.0029	8		518.43	5
520	Chromium, Total (Cr) (ppm)	0186	2.270	0.0200	4.751	1.702	0.1358	10	-1.46	520.52	0
520	Chromium, Total (Cr) (ppm)	0171	2.505	0.0100	4.751	1.702	0.1358	10	-1.32	520.41	0
520	Chromium, Total (Cr) (ppm)	0407	3.168	0.3418	4.751	1.702	0.1358	10	-0.93	520.41	0
520	Chromium, Total (Cr) (ppm)	0407	5.037	0.0939	4.751	1.702	0.1358	10	0.17	520.43	0
520	Chromium, Total (Cr) (ppm)	0407	5.422	0.1162	4.751	1.702	0.1358	10	0.39	520.53	0
520	Chromium, Total (Cr) (ppm)	0164	5.512	0.3660	4.751	1.702	0.1358	10	0.45	520.53	0
520	Chromium, Total (Cr) (ppm)	0553	5.625	0.1300	4.751	1.702	0.1358	10	0.51	520.53	0
520	Chromium, Total (Cr) (ppm)	0510	5.860	0.0200	4.751	1.702	0.1358	10	0.65	520.43	0
520	Chromium, Total (Cr) (ppm)	0045	5.890	0.2600	4.751	1.702	0.1358	10	0.67	520.42	0
520	Chromium, Total (Cr) (ppm)	0682	6.220	0.0000	4.751	1.702	0.1358	10	0.86	520.43	0
520	Chromium, Total (Cr) (ppm)	0297	5.500	1.000	4.751	1.702	0.1358	10	0.44	520.43	1
520	Chromium, Total (Cr) (ppm)	0656	< 0.1565		4.751	1.702	0.1358	10		520.31	5
526	Lead (ppm)	0407	0.0713	0.0016	0.0838	0.0109	0.0053	6	-1.15	526.53	0
526	Lead (ppm)	0227	0.0785	0.0010	0.0838	0.0109	0.0053	6	-0.48	526.53	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
526	Lead (ppm)	2012	0.0825	0.0010	0.0838	0.0109	0.0053	6	-0.12	526.53	0
526	Lead (ppm)	0553	0.0836	0.0011	0.0838	0.0109	0.0053	6	-0.02	526.53	0
526	Lead (ppm)	0407	0.0867	0.0068	0.0838	0.0109	0.0053	6	0.27	526.41	0
526	Lead (ppm)	0186	0.1100	0.0200	0.0838	0.0109	0.0053	6	2.41	526.52	0
526	Lead (ppm)	0407	0.2335	0.1148	0.0838	0.0109	0.0053	6	13.78	526.43	1
526	Lead (ppm)	0910	< 0.1		0.0838	0.0109	0.0053	6		526.52	5
526	Lead (ppm)	0682	< 10		0.0838	0.0109	0.0053	6		526.43	5
529	Mercury (ppb)	2012	0.9485	0.0070				2		529.99	0
529	Mercury (ppb)	0407	1.141	0.0214				2		529.99	0
529	Mercury (ppb)	0227	< 10					2		529.99	5
529	Mercury (ppb)	0682	< 10000					2		529.99	5
539	Nickel (ppm)	0186	1.840	0.0400	3.252	0.9836	0.0412	7	-1.44	539.52	0
539	Nickel (ppm)	0171	2.715	0.0100	3.252	0.9836	0.0412	7	-0.55	539.41	0
539	Nickel (ppm)	0407	3.023	0.0688	3.252	0.9836	0.0412	7	-0.23	539.41	0
539	Nickel (ppm)	0407	3.104	0.0594	3.252	0.9836	0.0412	7	-0.15	539.53	0
539	Nickel (ppm)	0553	3.625	0.1100	3.252	0.9836	0.0412	7	0.38	539.53	0
539	Nickel (ppm)	0297	4.000	0.0000	3.252	0.9836	0.0412	7	0.76	539.43	0
539	Nickel (ppm)	0682	4.460	0.0000	3.252	0.9836	0.0412	7	1.23	539.43	0
539	Nickel (ppm)	0407	3.103	0.8896	3.252	0.9836	0.0412	7	-0.15	539.43	1
710	Lauric Acid (12:0) (% (w/w))	0226	< 0.005					0		710.99	5
710	Lauric Acid (12:0) (% (w/w))	0676	< 0.01					0		710.99	5
714	Myristic Acid (14:0) (% (w/w))	0226	0.0103	0.0003				1		714.99	0
716	Palmitic Acid (16:0) (% (w/w))	0226	0.7396	0.0015				1		716.99	0
718	Palmitoleic Acid (9c-16:1) (% (w/w))	0676	0.0100	0.0000				2		718.99	0
718	Palmitoleic Acid (9c-16:1) (% (w/w))	0226	0.0115	0.0001				2		718.99	0
722	Stearic Acid (18:0) (% (w/w))	0226	0.1073	0.0020				1		722.99	0
724	Oleic Acid (9c-18:1) (% (w/w))	0226	0.9404	0.0087				1		724.99	0
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0676	2.114	0.0380				2		726.99	0
726	Linoleic Acid (9c,12c-18:2) (% (w/w))	0226	2.165	0.0100				2		726.99	0
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v))	0676	0.1135	0.0030				2		728.99	0
728	alpha-Linolenic Acid (9c,12c,15c-18:3) (% (w/v))	0226	0.1167	0.0010				2		728.99	0
730	Arachidic Acid (20:0) (% (w/w))	0226	0.0139	0.0004				1		730.99	0
732	Gondoic Acid (11c-20:1) (% (w/w))	0226	0.0187	0.0010				1		732.99	0
736	Arachidonic Acid (5c,8c,11c,14c-20:4) (% (w/v))	0676	0.0000	0.0000				0		736.99	4
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c) (% (w/v))	0676	0.0000	0.0000				0		740.99	4
740	Eicosapentaenoic Acid EPA (5c,8c,11c,14c,17c) (% (w/v))	0226	< 0.005					0		740.99	5
742	Behenic Acid (22:0) (% (w/w))	0226	0.0101	0.0004				1		742.99	0
744	Erucic Acid (13c-22:1) (% (w/w))	0226	< 0.005					0		744.99	5
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c) (% (w/v))	0676	0.0000	0.0000				0		746.99	4
746	Docosapentaenoic Acid n-3 DPA (7c,10c,13c,16c,19c) (% (w/v))	0226	< 0.005					0		746.99	5
748	Lignoceric Acid (24:0) (% (w/w))	0226	0.0107	0.0008				1		748.99	0
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c,22c) (% (w/v))	0676	0.0000	0.0000				0		750.99	4

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Robust SD	R-bar	# Tests			
750	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16	0226	< 0.005						0	750.99	5
752	Nervonic Acid (24:1) isomers (% (w/w))	0226	< 0.005						0	752.99	5
754	Total n-3 Polyunsaturated (Omega-3) Fatty Aci	0676	0.1100	0.0000					1	754.99	0
756	Total n-6 Polyunsaturated (Omega-6) Fatty Aci	0676	2.120	0.0400					1	756.99	0
772	Total Fatty Acids (% (w/w))	0226	4.150	0.0000					1	772.99	0

Notes: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2 and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = Rejected for duplicates too far apart, 2 = Rejected as outlier, 5 = A reporting limit and 4 = zeros submitted as values. Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). Flag 3 indicates not used in statistics.