




AAFCO[®]
Association of American Feed Control Officials



Mycotoxin Contaminants Scheme

Dairy Feed

Test Material Code # 202262

Labs Reporting: 63

Analytes Reported 15

Issue Date : 05/31/2022

Analyte Proficiency Testing Report

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
600	Total Aflatoxin (ppb)	0520	2.800	0.8000	20.83	7.065	1.225	48	-2.55	600.14	0
600	Total Aflatoxin (ppb)	0014	12.25	1.700	20.83	7.065	1.225	48	-1.21	600.01	0
600	Total Aflatoxin (ppb)	0202	12.65	1.100	20.83	7.065	1.225	48	-1.16	600.12	0
600	Total Aflatoxin (ppb)	2098	13.22	0.0400	20.83	7.065	1.225	48	-1.08	600.01	0
600	Total Aflatoxin (ppb)	0066	13.40	0.2000	20.83	7.065	1.225	48	-1.05	600.13	0
600	Total Aflatoxin (ppb)	0227	13.50	1.000	20.83	7.065	1.225	48	-1.04	600.01	0
600	Total Aflatoxin (ppb)	2369	13.52	0.6680	20.83	7.065	1.225	48	-1.03	600.14	0
600	Total Aflatoxin (ppb)	2369	13.52	0.6680	20.83	7.065	1.225	48	-1.03	600.20	0
600	Total Aflatoxin (ppb)	2196	14.26	0.0000	20.83	7.065	1.225	48	-0.93	600.20	0
600	Total Aflatoxin (ppb)	0723	14.47	0.0200	20.83	7.065	1.225	48	-0.90	600.20	0
600	Total Aflatoxin (ppb)	2147	15.00	8.000	20.83	7.065	1.225	48	-0.83	600.04	0
600	Total Aflatoxin (ppb)	0297	15.30	0.2000	20.83	7.065	1.225	48	-0.78	600.01	0
600	Total Aflatoxin (ppb)	0511	15.30	0.4000	20.83	7.065	1.225	48	-0.78	600.01	0
600	Total Aflatoxin (ppb)	0015	16.53	0.4000	20.83	7.065	1.225	48	-0.61	600.99	0
600	Total Aflatoxin (ppb)	0027	17.25	1.196	20.83	7.065	1.225	48	-0.51	600.21	0
600	Total Aflatoxin (ppb)	2033	17.30	1.600	20.83	7.065	1.225	48	-0.50	600.24	0
600	Total Aflatoxin (ppb)	0034	17.80	0.8000	20.83	7.065	1.225	48	-0.43	600.01	0
600	Total Aflatoxin (ppb)	0529	17.90	0.6000	20.83	7.065	1.225	48	-0.42	600.01	0
600	Total Aflatoxin (ppb)	2269	18.00	1.600	20.83	7.065	1.225	48	-0.40	600.98	0
600	Total Aflatoxin (ppb)	0876	18.60	4.200	20.83	7.065	1.225	48	-0.32	600.01	0
600	Total Aflatoxin (ppb)	2113	19.00	4.000	20.83	7.065	1.225	48	-0.26	600.01	0
600	Total Aflatoxin (ppb)	2199	19.00	0.0000	20.83	7.065	1.225	48	-0.26	600.14	0
600	Total Aflatoxin (ppb)	2096	19.05	0.7000	20.83	7.065	1.225	48	-0.25	600.01	0
600	Total Aflatoxin (ppb)	2387	19.20	3.200	20.83	7.065	1.225	48	-0.23	600.98	0
600	Total Aflatoxin (ppb)	0425	19.40	1.200	20.83	7.065	1.225	48	-0.20	600.01	0
600	Total Aflatoxin (ppb)	2022	19.54	0.3600	20.83	7.065	1.225	48	-0.18	600.08	0
600	Total Aflatoxin (ppb)	2226	19.65	0.2830	20.83	7.065	1.225	48	-0.17	600.12	0
600	Total Aflatoxin (ppb)	2268	20.50	1.000	20.83	7.065	1.225	48	-0.05	600.14	0
600	Total Aflatoxin (ppb)	0629	21.65	0.7000	20.83	7.065	1.225	48	0.12	600.20	0
600	Total Aflatoxin (ppb)	0004	23.00	2.000	20.83	7.065	1.225	48	0.31	600.14	0
600	Total Aflatoxin (ppb)	0142	24.05	0.1000	20.83	7.065	1.225	48	0.46	600.98	0
600	Total Aflatoxin (ppb)	0723	24.09	2.160	20.83	7.065	1.225	48	0.46	600.08	0
600	Total Aflatoxin (ppb)	0042	24.75	0.3000	20.83	7.065	1.225	48	0.55	600.20	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
600	Total Aflatoxin (ppb)	0682	24.80	0.0000	20.83	7.065	1.225	48	0.56	600.24	0
600	Total Aflatoxin (ppb)	0353	25.50	1.000	20.83	7.065	1.225	48	0.66	600.24	0
600	Total Aflatoxin (ppb)	2192	25.85	1.500	20.83	7.065	1.225	48	0.71	600.20	0
600	Total Aflatoxin (ppb)	2196	26.42	0.0000	20.83	7.065	1.225	48	0.79	600.08	0
600	Total Aflatoxin (ppb)	0098	27.37	0.4300	20.83	7.065	1.225	48	0.92	600.22	0
600	Total Aflatoxin (ppb)	0010	27.40	1.800	20.83	7.065	1.225	48	0.93	600.20	0
600	Total Aflatoxin (ppb)	0051	27.50	5.600	20.83	7.065	1.225	48	0.94	600.20	0
600	Total Aflatoxin (ppb)	0563	28.48	1.930	20.83	7.065	1.225	48	1.08	600.20	0
600	Total Aflatoxin (ppb)	2129	29.05	0.3000	20.83	7.065	1.225	48	1.16	600.24	0
600	Total Aflatoxin (ppb)	0035	29.31	0.1000	20.83	7.065	1.225	48	1.20	600.21	0
600	Total Aflatoxin (ppb)	0218	30.05	0.1000	20.83	7.065	1.225	48	1.30	600.24	0
600	Total Aflatoxin (ppb)	0407	31.94	2.630	20.83	7.065	1.225	48	1.57	600.13	0
600	Total Aflatoxin (ppb)	0674	32.18	0.0000	20.83	7.065	1.225	48	1.61	600.20	0
600	Total Aflatoxin (ppb)	2352	32.53	0.5700	20.83	7.065	1.225	48	1.65	600.23	0
600	Total Aflatoxin (ppb)	0035	39.83	1.640	20.83	7.065	1.225	48	2.69	600.13	0
600	Total Aflatoxin (ppb)	2302	85.20	2.710	20.83	7.065	1.225	48	9.11	600.99	2
601	Aflatoxin B1 (ppb)	2355	5.400	0.0000	19.01	6.462	1.430	36	-2.11	601.98	0
601	Aflatoxin B1 (ppb)	0009	5.460	0.2200	19.01	6.462	1.430	36	-2.10	601.23	0
601	Aflatoxin B1 (ppb)	2369	10.59	0.4460	19.01	6.462	1.430	36	-1.30	601.20	0
601	Aflatoxin B1 (ppb)	2369	10.59	0.4460	19.01	6.462	1.430	36	-1.30	601.24	0
601	Aflatoxin B1 (ppb)	2196	11.31	0.0000	19.01	6.462	1.430	36	-1.19	601.20	0
601	Aflatoxin B1 (ppb)	0723	11.89	0.0400	19.01	6.462	1.430	36	-1.10	601.20	0
601	Aflatoxin B1 (ppb)	0425	12.65	0.7000	19.01	6.462	1.430	36	-0.98	601.92	0
601	Aflatoxin B1 (ppb)	2053	13.34	1.210	19.01	6.462	1.430	36	-0.88	601.23	0
601	Aflatoxin B1 (ppb)	0027	14.10	0.7420	19.01	6.462	1.430	36	-0.76	601.21	0
601	Aflatoxin B1 (ppb)	2033	14.20	1.400	19.01	6.462	1.430	36	-0.74	601.23	0
601	Aflatoxin B1 (ppb)	0870	14.44	0.1600	19.01	6.462	1.430	36	-0.71	601.23	0
601	Aflatoxin B1 (ppb)	0028	14.93	3.810	19.01	6.462	1.430	36	-0.63	601.23	0
601	Aflatoxin B1 (ppb)	2192	15.60	1.800	19.01	6.462	1.430	36	-0.53	601.20	0
601	Aflatoxin B1 (ppb)	0160	16.05	2.300	19.01	6.462	1.430	36	-0.46	601.23	0
601	Aflatoxin B1 (ppb)	0918	16.25	0.5000	19.01	6.462	1.430	36	-0.43	601.23	0
601	Aflatoxin B1 (ppb)	2022	17.73	0.4600	19.01	6.462	1.430	36	-0.20	601.91	0
601	Aflatoxin B1 (ppb)	0629	17.85	0.7000	19.01	6.462	1.430	36	-0.18	601.20	0
601	Aflatoxin B1 (ppb)	2269	18.45	2.500	19.01	6.462	1.430	36	-0.09	601.98	0
601	Aflatoxin B1 (ppb)	0682	19.70	0.0000	19.01	6.462	1.430	36	0.11	601.23	0
601	Aflatoxin B1 (ppb)	0353	20.50	1.000	19.01	6.462	1.430	36	0.23	601.23	0
601	Aflatoxin B1 (ppb)	2140	20.50	5.000	19.01	6.462	1.430	36	0.23	601.23	0
601	Aflatoxin B1 (ppb)	2327	21.45	0.1000	19.01	6.462	1.430	36	0.38	601.23	0
601	Aflatoxin B1 (ppb)	0098	22.21	0.0600	19.01	6.462	1.430	36	0.50	601.24	0
601	Aflatoxin B1 (ppb)	0010	22.40	1.800	19.01	6.462	1.430	36	0.53	601.20	0
601	Aflatoxin B1 (ppb)	0913	22.70	1.000	19.01	6.462	1.430	36	0.57	601.23	0
601	Aflatoxin B1 (ppb)	0910	23.40	0.8000	19.01	6.462	1.430	36	0.68	601.21	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
601	Aflatoxin B1 (ppb)	0218	23.60	0.6000	19.01	6.462	1.430	36	0.71	601.23	0
601	Aflatoxin B1 (ppb)	2136	24.15	9.300	19.01	6.462	1.430	36	0.80	601.23	0
601	Aflatoxin B1 (ppb)	0563	24.51	1.740	19.01	6.462	1.430	36	0.85	601.20	0
601	Aflatoxin B1 (ppb)	0035	24.55	0.2200	19.01	6.462	1.430	36	0.86	601.21	0
601	Aflatoxin B1 (ppb)	2352	25.92	0.5400	19.01	6.462	1.430	36	1.07	601.23	0
601	Aflatoxin B1 (ppb)	0242	26.00	6.000	19.01	6.462	1.430	36	1.08	601.23	0
601	Aflatoxin B1 (ppb)	0644	28.13	0.0000	19.01	6.462	1.430	36	1.41	601.23	0
601	Aflatoxin B1 (ppb)	0553	28.53	3.270	19.01	6.462	1.430	36	1.47	601.23	0
601	Aflatoxin B1 (ppb)	0043	30.65	2.620	19.01	6.462	1.430	36	1.80	601.23	0
601	Aflatoxin B1 (ppb)	2302	65.52	0.0000	19.01	6.462	1.430	36	7.20	601.99	0
602	Aflatoxin B2 (ppb)	0009	1.380	0.0400	4.377	1.548	0.2151	28	-1.94	602.23	0
602	Aflatoxin B2 (ppb)	0723	2.580	0.0200	4.377	1.548	0.2151	28	-1.16	602.20	0
602	Aflatoxin B2 (ppb)	2369	2.928	0.2220	4.377	1.548	0.2151	28	-0.94	602.20	0
602	Aflatoxin B2 (ppb)	2369	2.928	0.2220	4.377	1.548	0.2151	28	-0.94	602.24	0
602	Aflatoxin B2 (ppb)	2196	2.950	0.0000	4.377	1.548	0.2151	28	-0.92	602.20	0
602	Aflatoxin B2 (ppb)	2192	3.130	0.2400	4.377	1.548	0.2151	28	-0.81	602.20	0
602	Aflatoxin B2 (ppb)	0027	3.145	0.4540	4.377	1.548	0.2151	28	-0.80	602.21	0
602	Aflatoxin B2 (ppb)	0918	3.215	0.1100	4.377	1.548	0.2151	28	-0.75	602.23	0
602	Aflatoxin B2 (ppb)	2053	3.520	0.4600	4.377	1.548	0.2151	28	-0.55	602.23	0
602	Aflatoxin B2 (ppb)	0160	3.690	0.1400	4.377	1.548	0.2151	28	-0.44	602.23	0
602	Aflatoxin B2 (ppb)	2327	3.705	0.2700	4.377	1.548	0.2151	28	-0.43	602.23	0
602	Aflatoxin B2 (ppb)	0629	3.800	0.0000	4.377	1.548	0.2151	28	-0.37	602.20	0
602	Aflatoxin B2 (ppb)	0563	3.965	0.1900	4.377	1.548	0.2151	28	-0.27	602.20	0
602	Aflatoxin B2 (ppb)	0870	4.020	0.2000	4.377	1.548	0.2151	28	-0.23	602.23	0
602	Aflatoxin B2 (ppb)	0682	4.200	0.0000	4.377	1.548	0.2151	28	-0.11	602.23	0
602	Aflatoxin B2 (ppb)	0910	4.515	0.0700	4.377	1.548	0.2151	28	0.09	602.21	0
602	Aflatoxin B2 (ppb)	0035	4.760	0.3200	4.377	1.548	0.2151	28	0.25	602.21	0
602	Aflatoxin B2 (ppb)	0010	4.950	0.1000	4.377	1.548	0.2151	28	0.37	602.20	0
602	Aflatoxin B2 (ppb)	0353	5.000	0.0000	4.377	1.548	0.2151	28	0.40	602.23	0
602	Aflatoxin B2 (ppb)	0553	5.045	0.5900	4.377	1.548	0.2151	28	0.43	602.23	0
602	Aflatoxin B2 (ppb)	0913	5.100	0.2000	4.377	1.548	0.2151	28	0.47	602.23	0
602	Aflatoxin B2 (ppb)	0098	5.155	0.4900	4.377	1.548	0.2151	28	0.50	602.24	0
602	Aflatoxin B2 (ppb)	2136	5.445	0.4700	4.377	1.548	0.2151	28	0.69	602.23	0
602	Aflatoxin B2 (ppb)	0043	6.288	0.3840	4.377	1.548	0.2151	28	1.23	602.23	0
602	Aflatoxin B2 (ppb)	0218	6.500	0.8000	4.377	1.548	0.2151	28	1.37	602.23	0
602	Aflatoxin B2 (ppb)	2352	6.605	0.0300	4.377	1.548	0.2151	28	1.44	602.23	0
602	Aflatoxin B2 (ppb)	0644	10.60	0.0000	4.377	1.548	0.2151	28	4.02	602.23	0
602	Aflatoxin B2 (ppb)	2140	14.00	0.0000	4.377	1.548	0.2151	28	6.21	602.23	0
602	Aflatoxin B2 (ppb)	0028	< 0.295		4.377	1.548	0.2151	28		602.23	5
602	Aflatoxin B2 (ppb)	0242	< 1		4.377	1.548	0.2151	28		602.23	5
602	Aflatoxin B2 (ppb)	2033	< 5		4.377	1.548	0.2151	28		602.23	5
603	Aflatoxin G1 (ppb)	0910	0.4220	0.0880	0.9658	0.4208	0.0220	4		603.21	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
603	Aflatoxin G1 (ppb)	0644	0.5410	0.0000	0.9658	0.4208	0.0220	4	603.23	0	
603	Aflatoxin G1 (ppb)	0682	0.9000	0.0000	0.9658	0.4208	0.0220	4	603.23	0	
603	Aflatoxin G1 (ppb)	0242	2.000	0.0000	0.9658	0.4208	0.0220	4	603.23	0	
603	Aflatoxin G1 (ppb)	2302	19.63	2.610	0.9658	0.4208	0.0220	4	603.99	2	
603	Aflatoxin G1 (ppb)	0043	< 0.1062		0.9658	0.4208	0.0220	4	603.23	5	
603	Aflatoxin G1 (ppb)	0028	< 0.176		0.9658	0.4208	0.0220	4	603.23	5	
603	Aflatoxin G1 (ppb)	0629	< 0.25		0.9658	0.4208	0.0220	4	603.20	5	
603	Aflatoxin G1 (ppb)	2053	< 0.32		0.9658	0.4208	0.0220	4	603.23	5	
603	Aflatoxin G1 (ppb)	2352	< 0.5		0.9658	0.4208	0.0220	4	603.23	5	
603	Aflatoxin G1 (ppb)	0870	< 0.6		0.9658	0.4208	0.0220	4	603.23	5	
603	Aflatoxin G1 (ppb)	2136	< 0.83		0.9658	0.4208	0.0220	4	603.23	5	
603	Aflatoxin G1 (ppb)	0160	< 1		0.9658	0.4208	0.0220	4	603.23	5	
603	Aflatoxin G1 (ppb)	0218	< 1		0.9658	0.4208	0.0220	4	603.23	5	
603	Aflatoxin G1 (ppb)	0553	< 1		0.9658	0.4208	0.0220	4	603.23	5	
603	Aflatoxin G1 (ppb)	0913	< 1		0.9658	0.4208	0.0220	4	603.23	5	
603	Aflatoxin G1 (ppb)	0918	< 1		0.9658	0.4208	0.0220	4	603.23	5	
603	Aflatoxin G1 (ppb)	2327	< 1.1		0.9658	0.4208	0.0220	4	603.23	5	
603	Aflatoxin G1 (ppb)	2033	< 5		0.9658	0.4208	0.0220	4	603.23	5	
603	Aflatoxin G1 (ppb)	0098	< 15		0.9658	0.4208	0.0220	4	603.24	5	
603	Aflatoxin G1 (ppb)	2140	< 20		0.9658	0.4208	0.0220	4	603.23	5	
604	Aflatoxin G2 (ppb)	0910	0.1158	0.0485				2	604.21	0	
604	Aflatoxin G2 (ppb)	0644	1.010	0.0000				2	604.23	0	
604	Aflatoxin G2 (ppb)	2369	0.2420	0.0660				2	604.20	1	
604	Aflatoxin G2 (ppb)	2369	0.2420	0.0660				2	604.24	1	
604	Aflatoxin G2 (ppb)	2389	18.50	0.0000				2	604.98	2	
604	Aflatoxin G2 (ppb)	0629	< 0.25					2	604.20	5	
604	Aflatoxin G2 (ppb)	0028	< 0.332					2	604.23	5	
604	Aflatoxin G2 (ppb)	2053	< 0.49					2	604.23	5	
604	Aflatoxin G2 (ppb)	0682	< 0.5					2	604.23	5	
604	Aflatoxin G2 (ppb)	0870	< 0.6					2	604.23	5	
604	Aflatoxin G2 (ppb)	2136	< 0.83					2	604.23	5	
604	Aflatoxin G2 (ppb)	0160	< 1					2	604.23	5	
604	Aflatoxin G2 (ppb)	0218	< 1					2	604.23	5	
604	Aflatoxin G2 (ppb)	0242	< 1					2	604.23	5	
604	Aflatoxin G2 (ppb)	0553	< 1					2	604.23	5	
604	Aflatoxin G2 (ppb)	0913	< 1					2	604.23	5	
604	Aflatoxin G2 (ppb)	0918	< 1					2	604.23	5	
604	Aflatoxin G2 (ppb)	2327	< 1.6					2	604.23	5	
604	Aflatoxin G2 (ppb)	2033	< 5					2	604.23	5	
604	Aflatoxin G2 (ppb)	0098	< 5					2	604.24	5	
604	Aflatoxin G2 (ppb)	2140	< 20					2	604.23	5	
610	Deoxynivalenol (ppb)	0014	0.0038	0.0004	3,194	945.3	232.6	56	-3.38	610.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
610	Deoxynivalenol (ppb)	2369	1.734	0.0620	3,194	945.3	232.6	56	-3.38	610.20	0
610	Deoxynivalenol (ppb)	0425	2.900	0.0000	3,194	945.3	232.6	56	-3.38	610.01	0
610	Deoxynivalenol (ppb)	0910	3.870	0.2000	3,194	945.3	232.6	56	-3.38	610.20	0
610	Deoxynivalenol (ppb)	2389	4.250	0.3000	3,194	945.3	232.6	56	-3.37	610.98	0
610	Deoxynivalenol (ppb)	2053	1,665	266.0	3,194	945.3	232.6	56	-1.62	610.23	0
610	Deoxynivalenol (ppb)	0015	2,042	131.6	3,194	945.3	232.6	56	-1.22	610.99	0
610	Deoxynivalenol (ppb)	0027	2,150	100.0	3,194	945.3	232.6	56	-1.10	610.01	0
610	Deoxynivalenol (ppb)	0098	2,150	500.0	3,194	945.3	232.6	56	-1.10	610.20	0
610	Deoxynivalenol (ppb)	0042	2,300	200.0	3,194	945.3	232.6	56	-0.95	610.01	0
610	Deoxynivalenol (ppb)	2196	2,389	0.0000	3,194	945.3	232.6	56	-0.85	610.05	0
610	Deoxynivalenol (ppb)	0066	2,475	50.00	3,194	945.3	232.6	56	-0.76	610.08	0
610	Deoxynivalenol (ppb)	2136	2,475	110.0	3,194	945.3	232.6	56	-0.76	610.23	0
610	Deoxynivalenol (ppb)	0511	2,500	200.0	3,194	945.3	232.6	56	-0.73	610.01	0
610	Deoxynivalenol (ppb)	0208	2,510	100.0	3,194	945.3	232.6	56	-0.72	610.01	0
610	Deoxynivalenol (ppb)	2268	2,607	14.00	3,194	945.3	232.6	56	-0.62	610.22	0
610	Deoxynivalenol (ppb)	0353	2,608	297.0	3,194	945.3	232.6	56	-0.62	610.23	0
610	Deoxynivalenol (ppb)	2196	2,641	0.0000	3,194	945.3	232.6	56	-0.59	610.20	0
610	Deoxynivalenol (ppb)	2113	2,650	100.0	3,194	945.3	232.6	56	-0.58	610.01	0
610	Deoxynivalenol (ppb)	0242	2,683	243.0	3,194	945.3	232.6	56	-0.54	610.23	0
610	Deoxynivalenol (ppb)	2269	3,015	630.0	3,194	945.3	232.6	56	-0.19	610.98	0
610	Deoxynivalenol (ppb)	2140	3,034	119.0	3,194	945.3	232.6	56	-0.17	610.23	0
610	Deoxynivalenol (ppb)	0028	3,184	231.5	3,194	945.3	232.6	56	-0.01	610.23	0
610	Deoxynivalenol (ppb)	0202	3,185	1,310	3,194	945.3	232.6	56	-0.01	610.06	0
610	Deoxynivalenol (ppb)	0529	3,200	200.0	3,194	945.3	232.6	56	0.01	610.01	0
610	Deoxynivalenol (ppb)	0674	3,220	0.0000	3,194	945.3	232.6	56	0.03	610.20	0
610	Deoxynivalenol (ppb)	2033	3,230	112.0	3,194	945.3	232.6	56	0.04	610.23	0
610	Deoxynivalenol (ppb)	0723	3,240	60.00	3,194	945.3	232.6	56	0.05	610.05	0
610	Deoxynivalenol (ppb)	0407	3,250	780.0	3,194	945.3	232.6	56	0.06	610.07	0
610	Deoxynivalenol (ppb)	2129	3,311	9.000	3,194	945.3	232.6	56	0.12	610.23	0
610	Deoxynivalenol (ppb)	2098	3,315	210.0	3,194	945.3	232.6	56	0.13	610.01	0
610	Deoxynivalenol (ppb)	0644	3,373	0.0000	3,194	945.3	232.6	56	0.19	610.23	0
610	Deoxynivalenol (ppb)	0553	3,394	126.7	3,194	945.3	232.6	56	0.21	610.23	0
610	Deoxynivalenol (ppb)	0051	3,411	38.00	3,194	945.3	232.6	56	0.23	610.20	0
610	Deoxynivalenol (ppb)	2355	3,450	700.0	3,194	945.3	232.6	56	0.27	610.98	0
610	Deoxynivalenol (ppb)	0227	3,500	200.0	3,194	945.3	232.6	56	0.32	610.01	0
610	Deoxynivalenol (ppb)	0010	3,538	707.0	3,194	945.3	232.6	56	0.36	610.23	0
610	Deoxynivalenol (ppb)	0009	3,594	209.6	3,194	945.3	232.6	56	0.42	610.23	0
610	Deoxynivalenol (ppb)	2022	3,635	10.00	3,194	945.3	232.6	56	0.47	610.05	0
610	Deoxynivalenol (ppb)	0682	3,710	0.0000	3,194	945.3	232.6	56	0.55	610.23	0
610	Deoxynivalenol (ppb)	0218	3,720	560.0	3,194	945.3	232.6	56	0.56	610.23	0
610	Deoxynivalenol (ppb)	0160	3,750	700.0	3,194	945.3	232.6	56	0.59	610.99	0
610	Deoxynivalenol (ppb)	0870	3,825	150.0	3,194	945.3	232.6	56	0.67	610.23	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
610	Deoxynivalenol (ppb)	0913	3,850	100.0	3,194	945.3	232.6	56	0.69	610.23	0
610	Deoxynivalenol (ppb)	2352	4,003	51.49	3,194	945.3	232.6	56	0.86	610.23	0
610	Deoxynivalenol (ppb)	0918	4,022	192.9	3,194	945.3	232.6	56	0.88	610.23	0
610	Deoxynivalenol (ppb)	0297	4,050	300.0	3,194	945.3	232.6	56	0.91	610.01	0
610	Deoxynivalenol (ppb)	0870	4,070	249.9	3,194	945.3	232.6	56	0.93	610.99	0
610	Deoxynivalenol (ppb)	2387	4,083	100.0	3,194	945.3	232.6	56	0.94	610.98	0
610	Deoxynivalenol (ppb)	2147	4,350	1,500	3,194	945.3	232.6	56	1.22	610.04	0
610	Deoxynivalenol (ppb)	2192	4,555	90.00	3,194	945.3	232.6	56	1.44	610.01	0
610	Deoxynivalenol (ppb)	0142	4,800	200.0	3,194	945.3	232.6	56	1.70	610.98	0
610	Deoxynivalenol (ppb)	2302	5,004	58.29	3,194	945.3	232.6	56	1.91	610.99	0
610	Deoxynivalenol (ppb)	0035	5,343	146.0	3,194	945.3	232.6	56	2.27	610.08	0
610	Deoxynivalenol (ppb)	2327	5,644	143.0	3,194	945.3	232.6	56	2.59	610.23	0
610	Deoxynivalenol (ppb)	0563	7,320	519.0	3,194	945.3	232.6	56	4.36	610.25	0
620	Total Fumonisin (ppb)	0014	0.0028	0.0001	2,782	826.3	250.2	34	-3.37	620.01	0
620	Total Fumonisin (ppb)	0425	2.950	0.3000	2,782	826.3	250.2	34	-3.36	620.01	0
620	Total Fumonisin (ppb)	0520	3.200	0.0000	2,782	826.3	250.2	34	-3.36	620.12	0
620	Total Fumonisin (ppb)	0098	850.0	100.0	2,782	826.3	250.2	34	-2.34	620.11	0
620	Total Fumonisin (ppb)	2369	1,766	42.04	2,782	826.3	250.2	34	-1.23	620.11	0
620	Total Fumonisin (ppb)	0015	1,970	146.4	2,782	826.3	250.2	34	-0.98	620.99	0
620	Total Fumonisin (ppb)	0674	2,020	0.0000	2,782	826.3	250.2	34	-0.92	620.07	0
620	Total Fumonisin (ppb)	2268	2,050	100.0	2,782	826.3	250.2	34	-0.89	620.11	0
620	Total Fumonisin (ppb)	2226	2,108	57.00	2,782	826.3	250.2	34	-0.82	620.09	0
620	Total Fumonisin (ppb)	0202	2,365	830.0	2,782	826.3	250.2	34	-0.50	620.09	0
620	Total Fumonisin (ppb)	0010	2,450	100.0	2,782	826.3	250.2	34	-0.40	620.23	0
620	Total Fumonisin (ppb)	0682	2,580	0.0000	2,782	826.3	250.2	34	-0.24	620.23	0
620	Total Fumonisin (ppb)	2192	2,590	740.0	2,782	826.3	250.2	34	-0.23	620.01	0
620	Total Fumonisin (ppb)	0297	2,600	0.0000	2,782	826.3	250.2	34	-0.22	620.01	0
620	Total Fumonisin (ppb)	2113	2,650	300.0	2,782	826.3	250.2	34	-0.16	620.01	0
620	Total Fumonisin (ppb)	0407	2,679	658.0	2,782	826.3	250.2	34	-0.12	620.09	0
620	Total Fumonisin (ppb)	2033	2,703	19.00	2,782	826.3	250.2	34	-0.10	620.23	0
620	Total Fumonisin (ppb)	0227	3,000	0.0000	2,782	826.3	250.2	34	0.26	620.01	0
620	Total Fumonisin (ppb)	0027	3,050	580.0	2,782	826.3	250.2	34	0.32	620.01	0
620	Total Fumonisin (ppb)	2098	3,090	100.0	2,782	826.3	250.2	34	0.37	620.02	0
620	Total Fumonisin (ppb)	0723	3,090	160.0	2,782	826.3	250.2	34	0.37	620.07	0
620	Total Fumonisin (ppb)	0042	3,100	400.0	2,782	826.3	250.2	34	0.38	620.01	0
620	Total Fumonisin (ppb)	0511	3,100	0.0000	2,782	826.3	250.2	34	0.38	620.01	0
620	Total Fumonisin (ppb)	0208	3,100	600.0	2,782	826.3	250.2	34	0.38	620.20	0
620	Total Fumonisin (ppb)	2147	3,250	1,300	2,782	826.3	250.2	34	0.57	620.04	0
620	Total Fumonisin (ppb)	0035	3,327	546.0	2,782	826.3	250.2	34	0.66	620.09	0
620	Total Fumonisin (ppb)	0529	3,500	400.0	2,782	826.3	250.2	34	0.87	620.02	0
620	Total Fumonisin (ppb)	2302	3,565	356.3	2,782	826.3	250.2	34	0.95	620.99	0
620	Total Fumonisin (ppb)	0142	3,600	200.0	2,782	826.3	250.2	34	0.99	620.98	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
620	Total Fumonisin (ppb)	0353	3,635	120.0	2,782	826.3	250.2	34	1.03	620.23	0
620	Total Fumonisin (ppb)	2129	3,760	461.0	2,782	826.3	250.2	34	1.18	620.23	0
620	Total Fumonisin (ppb)	2352	3,914	19.60	2,782	826.3	250.2	34	1.37	620.23	0
620	Total Fumonisin (ppb)	0066	4,025	150.0	2,782	826.3	250.2	34	1.50	620.09	0
620	Total Fumonisin (ppb)	2022	4,520	20.00	2,782	826.3	250.2	34	2.10	620.07	0
621	Fumonsin B1 (ppb)	0009	497.6	43.83	2,049	613.7	232.5	23	-2.53	621.23	0
621	Fumonsin B1 (ppb)	0098	850.0	100.0	2,049	613.7	232.5	23	-1.95	621.21	0
621	Fumonsin B1 (ppb)	0218	1,119	1,322	2,049	613.7	232.5	23	-1.52	621.23	0
621	Fumonsin B1 (ppb)	0913	1,150	100.0	2,049	613.7	232.5	23	-1.47	621.23	0
621	Fumonsin B1 (ppb)	2053	1,363	144.0	2,049	613.7	232.5	23	-1.12	621.23	0
621	Fumonsin B1 (ppb)	2369	1,413	26.92	2,049	613.7	232.5	23	-1.04	621.21	0
621	Fumonsin B1 (ppb)	0010	1,800	200.0	2,049	613.7	232.5	23	-0.41	621.23	0
621	Fumonsin B1 (ppb)	0563	1,828	215.0	2,049	613.7	232.5	23	-0.36	621.20	0
621	Fumonsin B1 (ppb)	0910	1,870	0.0000	2,049	613.7	232.5	23	-0.29	621.21	0
621	Fumonsin B1 (ppb)	2136	1,945	250.0	2,049	613.7	232.5	23	-0.17	621.23	0
621	Fumonsin B1 (ppb)	2033	1,975	10.00	2,049	613.7	232.5	23	-0.12	621.23	0
621	Fumonsin B1 (ppb)	0644	2,005	0.0000	2,049	613.7	232.5	23	-0.07	621.23	0
621	Fumonsin B1 (ppb)	0918	2,112	67.71	2,049	613.7	232.5	23	0.10	621.23	0
621	Fumonsin B1 (ppb)	0682	2,160	0.0000	2,049	613.7	232.5	23	0.18	621.23	0
621	Fumonsin B1 (ppb)	0870	2,240	80.00	2,049	613.7	232.5	23	0.31	621.23	0
621	Fumonsin B1 (ppb)	2140	2,325	53.00	2,049	613.7	232.5	23	0.45	621.23	0
621	Fumonsin B1 (ppb)	0043	2,477	287.0	2,049	613.7	232.5	23	0.70	621.23	0
621	Fumonsin B1 (ppb)	2327	2,634	218.0	2,049	613.7	232.5	23	0.95	621.23	0
621	Fumonsin B1 (ppb)	0160	2,750	300.0	2,049	613.7	232.5	23	1.14	621.23	0
621	Fumonsin B1 (ppb)	2352	2,844	32.07	2,049	613.7	232.5	23	1.29	621.23	0
621	Fumonsin B1 (ppb)	0353	2,923	13.00	2,049	613.7	232.5	23	1.42	621.23	0
621	Fumonsin B1 (ppb)	2355	3,850	100.0	2,049	613.7	232.5	23	2.93	621.98	0
621	Fumonsin B1 (ppb)	0553	6,605	1,785	2,049	613.7	232.5	23	7.42	621.23	0
622	Fumonisin B2 (ppb)	0009	181.1	58.40	569.2	176.5	34.97	19	-2.20	622.23	0
622	Fumonisin B2 (ppb)	0913	300.0	0.0000	569.2	176.5	34.97	19	-1.53	622.23	0
622	Fumonisin B2 (ppb)	2053	346.5	13.10	569.2	176.5	34.97	19	-1.26	622.23	0
622	Fumonisin B2 (ppb)	2369	353.4	15.12	569.2	176.5	34.97	19	-1.22	622.21	0
622	Fumonisin B2 (ppb)	0682	420.0	0.0000	569.2	176.5	34.97	19	-0.85	622.23	0
622	Fumonisin B2 (ppb)	2327	478.0	52.00	569.2	176.5	34.97	19	-0.52	622.23	0
622	Fumonisin B2 (ppb)	0010	500.0	0.0000	569.2	176.5	34.97	19	-0.39	622.23	0
622	Fumonisin B2 (ppb)	0910	509.0	0.0000	569.2	176.5	34.97	19	-0.34	622.21	0
622	Fumonisin B2 (ppb)	2033	511.5	9.000	569.2	176.5	34.97	19	-0.33	622.23	0
622	Fumonisin B2 (ppb)	2136	549.5	73.00	569.2	176.5	34.97	19	-0.11	622.23	0
622	Fumonisin B2 (ppb)	0870	550.0	20.00	569.2	176.5	34.97	19	-0.11	622.23	0
622	Fumonisin B2 (ppb)	2140	554.0	40.00	569.2	176.5	34.97	19	-0.09	622.23	0
622	Fumonisin B2 (ppb)	0644	596.8	0.0000	569.2	176.5	34.97	19	0.16	622.23	0
622	Fumonisin B2 (ppb)	0043	741.7	16.65	569.2	176.5	34.97	19	0.98	622.23	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
622	Fumonisin B2 (ppb)	0160	750.0	300.0	569.2	176.5	34.97	19	1.02	622.23	0
622	Fumonisin B2 (ppb)	0353	770.0	16.00	569.2	176.5	34.97	19	1.14	622.23	0
622	Fumonisin B2 (ppb)	2352	831.4	9.590	569.2	176.5	34.97	19	1.49	622.23	0
622	Fumonisin B2 (ppb)	0918	922.9	15.21	569.2	176.5	34.97	19	2.00	622.23	0
622	Fumonisin B2 (ppb)	0553	1,585	26.40	569.2	176.5	34.97	19	5.76	622.23	0
622	Fumonisin B2 (ppb)	0218	1,256	1,529	569.2	176.5	34.97	19	3.89	622.23	1
622	Fumonisin B2 (ppb)	0098	< 200		569.2	176.5	34.97	19		622.21	5
623	Fumonisin B3 (ppb)	0009	74.60	6.380	195.8	62.48	38.21	11	-1.94	623.23	0
623	Fumonisin B3 (ppb)	0160	100.0	0.0000	195.8	62.48	38.21	11	-1.53	623.23	0
623	Fumonisin B3 (ppb)	0010	150.0	100.0	195.8	62.48	38.21	11	-0.73	623.23	0
623	Fumonisin B3 (ppb)	0913	150.0	100.0	195.8	62.48	38.21	11	-0.73	623.23	0
623	Fumonisin B3 (ppb)	0353	163.5	71.00	195.8	62.48	38.21	11	-0.52	623.23	0
623	Fumonisin B3 (ppb)	2327	186.5	17.00	195.8	62.48	38.21	11	-0.15	623.23	0
623	Fumonisin B3 (ppb)	2033	216.0	0.0000	195.8	62.48	38.21	11	0.32	623.23	0
623	Fumonisin B3 (ppb)	2352	238.4	2.880	195.8	62.48	38.21	11	0.68	623.23	0
623	Fumonisin B3 (ppb)	2053	255.2	9.600	195.8	62.48	38.21	11	0.95	623.23	0
623	Fumonisin B3 (ppb)	0870	290.0	20.00	195.8	62.48	38.21	11	1.51	623.23	0
623	Fumonisin B3 (ppb)	0553	960.9	93.40	195.8	62.48	38.21	11	12.25	623.23	0
630	Ochratoxin A (ppb)	0425	17.50	1.000	106.8	34.66	11.57	38	-2.58	630.01	0
630	Ochratoxin A (ppb)	2098	24.60	0.8000	106.8	34.66	11.57	38	-2.37	630.01	0
630	Ochratoxin A (ppb)	0015	45.95	1.090	106.8	34.66	11.57	38	-1.76	630.99	0
630	Ochratoxin A (ppb)	2053	46.06	11.61	106.8	34.66	11.57	38	-1.75	630.23	0
630	Ochratoxin A (ppb)	0674	50.85	0.0000	106.8	34.66	11.57	38	-1.62	630.20	0
630	Ochratoxin A (ppb)	0520	55.00	6.000	106.8	34.66	11.57	38	-1.50	630.06	0
630	Ochratoxin A (ppb)	2226	64.00	14.20	106.8	34.66	11.57	38	-1.24	630.05	0
630	Ochratoxin A (ppb)	0553	72.74	5.310	106.8	34.66	11.57	38	-0.98	630.23	0
630	Ochratoxin A (ppb)	2136	76.25	15.30	106.8	34.66	11.57	38	-0.88	630.23	0
630	Ochratoxin A (ppb)	0407	76.73	15.26	106.8	34.66	11.57	38	-0.87	630.05	0
630	Ochratoxin A (ppb)	2192	82.00	12.40	106.8	34.66	11.57	38	-0.72	630.20	0
630	Ochratoxin A (ppb)	0723	84.03	0.0600	106.8	34.66	11.57	38	-0.66	630.20	0
630	Ochratoxin A (ppb)	2033	90.00	20.00	106.8	34.66	11.57	38	-0.49	630.23	0
630	Ochratoxin A (ppb)	0629	98.40	2.200	106.8	34.66	11.57	38	-0.24	630.20	0
630	Ochratoxin A (ppb)	2369	101.2	10.45	106.8	34.66	11.57	38	-0.16	630.06	0
630	Ochratoxin A (ppb)	2369	101.2	10.45	106.8	34.66	11.57	38	-0.16	630.20	0
630	Ochratoxin A (ppb)	2140	103.5	11.00	106.8	34.66	11.57	38	-0.10	630.23	0
630	Ochratoxin A (ppb)	0913	107.0	4.100	106.8	34.66	11.57	38	0.00	630.23	0
630	Ochratoxin A (ppb)	0218	109.0	2.000	106.8	34.66	11.57	38	0.06	630.23	0
630	Ochratoxin A (ppb)	0918	111.8	16.00	106.8	34.66	11.57	38	0.14	630.23	0
630	Ochratoxin A (ppb)	2327	112.0	4.000	106.8	34.66	11.57	38	0.15	630.23	0
630	Ochratoxin A (ppb)	0563	113.9	7.240	106.8	34.66	11.57	38	0.20	630.20	0
630	Ochratoxin A (ppb)	0066	114.2	9.370	106.8	34.66	11.57	38	0.21	630.05	0
630	Ochratoxin A (ppb)	0723	119.0	3.000	106.8	34.66	11.57	38	0.35	630.03	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
630	Ochratoxin A (ppb)	0910	119.0	28.00	106.8	34.66	11.57	38	0.35	630.21	0
630	Ochratoxin A (ppb)	2196	121.4	0.0000	106.8	34.66	11.57	38	0.42	630.20	0
630	Ochratoxin A (ppb)	0353	128.0	18.00	106.8	34.66	11.57	38	0.61	630.23	0
630	Ochratoxin A (ppb)	0682	132.0	0.0000	106.8	34.66	11.57	38	0.73	630.23	0
630	Ochratoxin A (ppb)	0870	132.9	23.80	106.8	34.66	11.57	38	0.75	630.23	0
630	Ochratoxin A (ppb)	2352	144.1	9.820	106.8	34.66	11.57	38	1.07	630.23	0
630	Ochratoxin A (ppb)	2129	149.0	14.00	106.8	34.66	11.57	38	1.22	630.23	0
630	Ochratoxin A (ppb)	0644	150.7	0.0000	106.8	34.66	11.57	38	1.27	630.23	0
630	Ochratoxin A (ppb)	2022	152.4	0.6800	106.8	34.66	11.57	38	1.31	630.03	0
630	Ochratoxin A (ppb)	0051	153.0	56.00	106.8	34.66	11.57	38	1.33	630.20	0
630	Ochratoxin A (ppb)	2302	156.7	9.500	106.8	34.66	11.57	38	1.44	630.99	0
630	Ochratoxin A (ppb)	0098	158.5	31.00	106.8	34.66	11.57	38	1.49	630.20	0
630	Ochratoxin A (ppb)	0160	175.0	30.00	106.8	34.66	11.57	38	1.97	630.23	0
630	Ochratoxin A (ppb)	0529	442.0	36.00	106.8	34.66	11.57	38	9.67	630.01	0
640	T-2 Toxin (ppb)	2053	25.83	5.740	45.88	15.23	2.942	19	-1.32	640.23	0
640	T-2 Toxin (ppb)	0682	34.40	0.0000	45.88	15.23	2.942	19	-0.75	640.23	0
640	T-2 Toxin (ppb)	0674	35.73	0.0000	45.88	15.23	2.942	19	-0.67	640.03	0
640	T-2 Toxin (ppb)	0913	36.45	4.900	45.88	15.23	2.942	19	-0.62	640.23	0
640	T-2 Toxin (ppb)	0910	36.70	6.000	45.88	15.23	2.942	19	-0.60	640.21	0
640	T-2 Toxin (ppb)	0015	38.08	1.470	45.88	15.23	2.942	19	-0.51	640.99	0
640	T-2 Toxin (ppb)	2140	39.00	0.0000	45.88	15.23	2.942	19	-0.45	640.23	0
640	T-2 Toxin (ppb)	0407	39.48	6.840	45.88	15.23	2.942	19	-0.42	640.05	0
640	T-2 Toxin (ppb)	2033	40.00	0.0000	45.88	15.23	2.942	19	-0.39	640.23	0
640	T-2 Toxin (ppb)	0218	42.50	7.000	45.88	15.23	2.942	19	-0.22	640.23	0
640	T-2 Toxin (ppb)	0644	43.69	0.0000	45.88	15.23	2.942	19	-0.14	640.23	0
640	T-2 Toxin (ppb)	2136	46.15	7.900	45.88	15.23	2.942	19	0.02	640.23	0
640	T-2 Toxin (ppb)	0918	49.70	3.170	45.88	15.23	2.942	19	0.25	640.23	0
640	T-2 Toxin (ppb)	2352	50.94	2.260	45.88	15.23	2.942	19	0.33	640.23	0
640	T-2 Toxin (ppb)	0870	54.56	1.080	45.88	15.23	2.942	19	0.57	640.23	0
640	T-2 Toxin (ppb)	0353	60.00	2.000	45.88	15.23	2.942	19	0.93	640.23	0
640	T-2 Toxin (ppb)	2129	76.80	3.400	45.88	15.23	2.942	19	2.03	640.23	0
640	T-2 Toxin (ppb)	0066	82.24	4.130	45.88	15.23	2.942	19	2.39	640.03	0
640	T-2 Toxin (ppb)	0297	108.6	0.0000	45.88	15.23	2.942	19	4.12	640.98	0
640	T-2 Toxin (ppb)	2269	191.2	42.40	45.88	15.23	2.942	19	9.54	640.98	1
640	T-2 Toxin (ppb)	0553	< 1		45.88	15.23	2.942	19		640.23	5
640	T-2 Toxin (ppb)	0563	< 20		45.88	15.23	2.942	19		640.25	5
640	T-2 Toxin (ppb)	0160	< 50		45.88	15.23	2.942	19		640.23	5
640	T-2 Toxin (ppb)	2327	< 200		45.88	15.23	2.942	19		640.23	5
641	HT-2 Toxin (ppb)	0910	27.40	6.400	54.87	18.13	2.707	11	-1.52	641.21	0
641	HT-2 Toxin (ppb)	0913	41.10	1.800	54.87	18.13	2.707	11	-0.76	641.23	0
641	HT-2 Toxin (ppb)	2033	45.45	2.500	54.87	18.13	2.707	11	-0.52	641.23	0
641	HT-2 Toxin (ppb)	2053	46.31	2.240	54.87	18.13	2.707	11	-0.47	641.23	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
641	HT-2 Toxin (ppb)	0870	50.90	0.7200	54.87	18.13	2.707	11	-0.22	641.23	0
641	HT-2 Toxin (ppb)	0353	56.00	0.0000	54.87	18.13	2.707	11	0.06	641.23	0
641	HT-2 Toxin (ppb)	2136	57.10	6.600	54.87	18.13	2.707	11	0.12	641.23	0
641	HT-2 Toxin (ppb)	0218	60.50	3.000	54.87	18.13	2.707	11	0.31	641.23	0
641	HT-2 Toxin (ppb)	0644	64.97	0.0000	54.87	18.13	2.707	11	0.56	641.23	0
641	HT-2 Toxin (ppb)	2352	71.51	4.520	54.87	18.13	2.707	11	0.92	641.23	0
641	HT-2 Toxin (ppb)	2140	124.0	2.000	54.87	18.13	2.707	11	3.81	641.23	0
641	HT-2 Toxin (ppb)	2327	< 200		54.87	18.13	2.707	11		641.23	5
642	T-2 Toxin + HT-2 Toxin (ppb)	0142	66.40	2.400	122.2	39.51	12.63	11	-1.41	642.98	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0407	75.53	10.19	122.2	39.51	12.63	11	-1.18	642.99	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0227	80.00	26.00	122.2	39.51	12.63	11	-1.07	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0425	101.5	19.00	122.2	39.51	12.63	11	-0.52	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0353	118.5	5.000	122.2	39.51	12.63	11	-0.09	642.04	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2302	123.6	1.070	122.2	39.51	12.63	11	0.04	642.99	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0529	137.0	2.000	122.2	39.51	12.63	11	0.37	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2022	142.9	0.7400	122.2	39.51	12.63	11	0.52	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2098	148.8	11.00	122.2	39.51	12.63	11	0.67	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2192	162.5	3.000	122.2	39.51	12.63	11	1.02	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2269	263.9	58.50	122.2	39.51	12.63	11	3.58	642.98	0
650	Zearalenone (ppb)	0520	12.50	1.000	545.4	169.3	36.10	50	-3.15	650.99	0
650	Zearalenone (ppb)	2098	127.6	10.80	545.4	169.3	36.10	50	-2.47	650.01	0
650	Zearalenone (ppb)	0242	129.5	37.00	545.4	169.3	36.10	50	-2.46	650.24	0
650	Zearalenone (ppb)	0553	168.8	90.70	545.4	169.3	36.10	50	-2.22	650.24	0
650	Zearalenone (ppb)	2053	265.7	35.70	545.4	169.3	36.10	50	-1.65	650.24	0
650	Zearalenone (ppb)	2147	295.0	50.00	545.4	169.3	36.10	50	-1.48	650.98	0
650	Zearalenone (ppb)	2113	315.0	10.00	545.4	169.3	36.10	50	-1.36	650.01	0
650	Zearalenone (ppb)	0066	345.3	24.60	545.4	169.3	36.10	50	-1.18	650.07	0
650	Zearalenone (ppb)	0035	389.8	42.45	545.4	169.3	36.10	50	-0.92	650.07	0
650	Zearalenone (ppb)	0407	390.3	4.140	545.4	169.3	36.10	50	-0.92	650.07	0
650	Zearalenone (ppb)	2268	400.0	0.0000	545.4	169.3	36.10	50	-0.86	650.99	0
650	Zearalenone (ppb)	2196	408.1	0.0000	545.4	169.3	36.10	50	-0.81	650.20	0
650	Zearalenone (ppb)	2226	419.9	27.65	545.4	169.3	36.10	50	-0.74	650.06	0
650	Zearalenone (ppb)	0529	424.6	49.20	545.4	169.3	36.10	50	-0.71	650.01	0
650	Zearalenone (ppb)	2302	436.5	0.0000	545.4	169.3	36.10	50	-0.64	650.99	0
650	Zearalenone (ppb)	2389	482.3	18.10	545.4	169.3	36.10	50	-0.37	650.98	0
650	Zearalenone (ppb)	2136	496.0	64.00	545.4	169.3	36.10	50	-0.29	650.24	0
650	Zearalenone (ppb)	0870	498.8	4.421	545.4	169.3	36.10	50	-0.28	650.01	0
650	Zearalenone (ppb)	2369	501.1	5.314	545.4	169.3	36.10	50	-0.26	650.20	0
650	Zearalenone (ppb)	2369	501.1	5.314	545.4	169.3	36.10	50	-0.26	650.21	0
650	Zearalenone (ppb)	0202	512.0	20.00	545.4	169.3	36.10	50	-0.20	650.06	0
650	Zearalenone (ppb)	0910	513.5	93.00	545.4	169.3	36.10	50	-0.19	650.20	0
650	Zearalenone (ppb)	0682	531.0	0.0000	545.4	169.3	36.10	50	-0.09	650.24	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
650	Zearalenone (ppb)	0425	532.1	30.90	545.4	169.3	36.10	50	-0.08	650.01	0
650	Zearalenone (ppb)	0723	538.2	0.0800	545.4	169.3	36.10	50	-0.04	650.20	0
650	Zearalenone (ppb)	0563	544.5	11.00	545.4	169.3	36.10	50	-0.01	650.20	0
650	Zearalenone (ppb)	0098	550.0	60.00	545.4	169.3	36.10	50	0.03	650.20	0
650	Zearalenone (ppb)	2022	561.4	3.230	545.4	169.3	36.10	50	0.09	650.04	0
650	Zearalenone (ppb)	2140	571.5	67.00	545.4	169.3	36.10	50	0.15	650.24	0
650	Zearalenone (ppb)	2192	572.0	36.00	545.4	169.3	36.10	50	0.16	650.20	0
650	Zearalenone (ppb)	0674	580.0	0.0000	545.4	169.3	36.10	50	0.20	650.20	0
650	Zearalenone (ppb)	0043	585.7	41.90	545.4	169.3	36.10	50	0.24	650.24	0
650	Zearalenone (ppb)	2033	595.0	50.00	545.4	169.3	36.10	50	0.29	650.24	0
650	Zearalenone (ppb)	0051	625.5	99.00	545.4	169.3	36.10	50	0.47	650.20	0
650	Zearalenone (ppb)	2352	643.3	1.120	545.4	169.3	36.10	50	0.58	650.24	0
650	Zearalenone (ppb)	0918	646.0	15.70	545.4	169.3	36.10	50	0.59	650.24	0
650	Zearalenone (ppb)	0723	651.9	19.04	545.4	169.3	36.10	50	0.63	650.04	0
650	Zearalenone (ppb)	0297	678.0	0.5000	545.4	169.3	36.10	50	0.78	650.01	0
650	Zearalenone (ppb)	0644	690.3	0.0000	545.4	169.3	36.10	50	0.86	650.24	0
650	Zearalenone (ppb)	0218	697.0	8.000	545.4	169.3	36.10	50	0.90	650.24	0
650	Zearalenone (ppb)	0142	711.8	55.40	545.4	169.3	36.10	50	0.98	650.98	0
650	Zearalenone (ppb)	0913	716.6	12.80	545.4	169.3	36.10	50	1.01	650.24	0
650	Zearalenone (ppb)	0010	735.5	103.0	545.4	169.3	36.10	50	1.12	650.24	0
650	Zearalenone (ppb)	2327	756.5	29.00	545.4	169.3	36.10	50	1.25	650.24	0
650	Zearalenone (ppb)	2269	763.1	83.70	545.4	169.3	36.10	50	1.29	650.98	0
650	Zearalenone (ppb)	2196	847.3	0.0000	545.4	169.3	36.10	50	1.78	650.04	0
650	Zearalenone (ppb)	2129	860.5	5.000	545.4	169.3	36.10	50	1.86	650.24	0
650	Zearalenone (ppb)	0353	884.5	73.00	545.4	169.3	36.10	50	2.00	650.23	0
650	Zearalenone (ppb)	0027	924.5	179.0	545.4	169.3	36.10	50	2.24	650.01	0
650	Zearalenone (ppb)	0160	966.5	227.0	545.4	169.3	36.10	50	2.49	650.24	0

Notes: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2 and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = Rejected for duplicates too far apart, 2 = Rejected as outlier, 5 = A reporting limit and 4 = zeros submitted as values. Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). Flag 3 indicates not used in statistics.