



AAFCO
Association of American Feed Control Officials



Mycotoxin Contaminants Scheme

Soybean Meal

Test Material Code # 202264

Labs Reporting: 68

Analytes Reported 15

Issue Date : 11/30/2022

Analyte Proficiency Testing Report

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
600	Total Aflatoxin (ppb)	0006	0.1415	0.0028	5.478	1.926	0.9206	47	-2.77	600.22	0
600	Total Aflatoxin (ppb)	2355	1.850	0.1000	5.478	1.926	0.9206	47	-1.88	600.98	0
600	Total Aflatoxin (ppb)	2326	2.230	0.3800	5.478	1.926	0.9206	47	-1.69	600.23	0
600	Total Aflatoxin (ppb)	0520	2.250	0.3000	5.478	1.926	0.9206	47	-1.68	600.14	0
600	Total Aflatoxin (ppb)	2268	2.550	0.7000	5.478	1.926	0.9206	47	-1.52	600.14	0
600	Total Aflatoxin (ppb)	2345	3.150	0.3000	5.478	1.926	0.9206	47	-1.21	600.14	0
600	Total Aflatoxin (ppb)	0596	3.950	0.5000	5.478	1.926	0.9206	47	-0.79	600.98	0
600	Total Aflatoxin (ppb)	0066	4.050	0.3000	5.478	1.926	0.9206	47	-0.74	600.13	0
600	Total Aflatoxin (ppb)	2033	4.150	0.7000	5.478	1.926	0.9206	47	-0.69	600.24	0
600	Total Aflatoxin (ppb)	0353	4.335	0.2100	5.478	1.926	0.9206	47	-0.59	600.24	0
600	Total Aflatoxin (ppb)	0723	4.395	1.010	5.478	1.926	0.9206	47	-0.56	600.20	0
600	Total Aflatoxin (ppb)	0723	4.520	0.0200	5.478	1.926	0.9206	47	-0.50	600.08	0
600	Total Aflatoxin (ppb)	0015	4.550	1.300	5.478	1.926	0.9206	47	-0.48	600.16	0
600	Total Aflatoxin (ppb)	0563	4.575	0.2700	5.478	1.926	0.9206	47	-0.47	600.20	0
600	Total Aflatoxin (ppb)	0511	4.600	0.4000	5.478	1.926	0.9206	47	-0.46	600.01	0
600	Total Aflatoxin (ppb)	2352	4.640	0.1800	5.478	1.926	0.9206	47	-0.43	600.23	0
600	Total Aflatoxin (ppb)	2369	4.650	0.4200	5.478	1.926	0.9206	47	-0.43	600.14	0
600	Total Aflatoxin (ppb)	2369	4.650	0.4200	5.478	1.926	0.9206	47	-0.43	600.20	0
600	Total Aflatoxin (ppb)	0629	4.795	0.5700	5.478	1.926	0.9206	47	-0.35	600.20	0
600	Total Aflatoxin (ppb)	0010	4.850	1.300	5.478	1.926	0.9206	47	-0.33	600.20	0
600	Total Aflatoxin (ppb)	0042	4.900	0.7600	5.478	1.926	0.9206	47	-0.30	600.20	0
600	Total Aflatoxin (ppb)	0682	5.000	0.0000	5.478	1.926	0.9206	47	-0.25	600.24	0
600	Total Aflatoxin (ppb)	2387	5.350	0.5000	5.478	1.926	0.9206	47	-0.07	600.98	0
600	Total Aflatoxin (ppb)	2022	5.590	0.4400	5.478	1.926	0.9206	47	0.06	600.08	0
600	Total Aflatoxin (ppb)	0297	5.750	0.1000	5.478	1.926	0.9206	47	0.14	600.01	0
600	Total Aflatoxin (ppb)	2192	5.800	0.2000	5.478	1.926	0.9206	47	0.17	600.20	0
600	Total Aflatoxin (ppb)	0529	5.850	1.300	5.478	1.926	0.9206	47	0.19	600.01	0
600	Total Aflatoxin (ppb)	0098	5.880	1.060	5.478	1.926	0.9206	47	0.21	600.22	0
600	Total Aflatoxin (ppb)	2022	5.895	0.1300	5.478	1.926	0.9206	47	0.22	600.01	0
600	Total Aflatoxin (ppb)	2411	5.975	0.9100	5.478	1.926	0.9206	47	0.26	600.24	0
600	Total Aflatoxin (ppb)	0034	6.100	0.8000	5.478	1.926	0.9206	47	0.32	600.01	0
600	Total Aflatoxin (ppb)	2113	6.100	1.800	5.478	1.926	0.9206	47	0.32	600.01	0
600	Total Aflatoxin (ppb)	0407	6.345	2.790	5.478	1.926	0.9206	47	0.45	600.13	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
600	Total Aflatoxin (ppb)	0218	6.350	0.3000	5.478	1.926	0.9206	47	0.45	600.24	0
600	Total Aflatoxin (ppb)	0876	6.450	3.700	5.478	1.926	0.9206	47	0.50	600.01	0
600	Total Aflatoxin (ppb)	0004	6.500	1.200	5.478	1.926	0.9206	47	0.53	600.14	0
600	Total Aflatoxin (ppb)	0918	6.865	1.190	5.478	1.926	0.9206	47	0.72	600.01	0
600	Total Aflatoxin (ppb)	2269	7.000	0.2200	5.478	1.926	0.9206	47	0.79	600.98	0
600	Total Aflatoxin (ppb)	2199	7.250	4.300	5.478	1.926	0.9206	47	0.92	600.14	0
600	Total Aflatoxin (ppb)	0033	7.445	4.810	5.478	1.926	0.9206	47	1.02	600.01	0
600	Total Aflatoxin (ppb)	2129	7.500	1.400	5.478	1.926	0.9206	47	1.05	600.24	0
600	Total Aflatoxin (ppb)	2096	7.650	2.100	5.478	1.926	0.9206	47	1.13	600.01	0
600	Total Aflatoxin (ppb)	0142	7.650	0.7000	5.478	1.926	0.9206	47	1.13	600.98	0
600	Total Aflatoxin (ppb)	0014	8.050	0.7000	5.478	1.926	0.9206	47	1.34	600.01	0
600	Total Aflatoxin (ppb)	0590	8.060	0.9400	5.478	1.926	0.9206	47	1.34	600.12	0
600	Total Aflatoxin (ppb)	0027	8.445	0.5100	5.478	1.926	0.9206	47	1.54	600.01	0
600	Total Aflatoxin (ppb)	2226	9.169	1.024	5.478	1.926	0.9206	47	1.92	600.12	0
600	Total Aflatoxin (ppb)	2246	23.95	0.4390	5.478	1.926	0.9206	47	9.59	600.08	2
600	Total Aflatoxin (ppb)	0032	0.0000	0.0000	5.478	1.926	0.9206	47		600.24	4
600	Total Aflatoxin (ppb)	0227	< 5		5.478	1.926	0.9206	47		600.01	5
600	Total Aflatoxin (ppb)	2098	< 5		5.478	1.926	0.9206	47		600.01	5
601	Aflatoxin B1 (ppb)	2326	2.225	0.3700	4.969	1.752	0.6259	31	-1.57	601.23	0
601	Aflatoxin B1 (ppb)	2246	2.410	0.0200	4.969	1.752	0.6259	31	-1.46	601.24	0
601	Aflatoxin B1 (ppb)	0913	3.750	0.5000	4.969	1.752	0.6259	31	-0.70	601.23	0
601	Aflatoxin B1 (ppb)	0870	4.000	0.4400	4.969	1.752	0.6259	31	-0.55	601.23	0
601	Aflatoxin B1 (ppb)	0353	4.040	0.2200	4.969	1.752	0.6259	31	-0.53	601.23	0
601	Aflatoxin B1 (ppb)	0723	4.060	0.9400	4.969	1.752	0.6259	31	-0.52	601.20	0
601	Aflatoxin B1 (ppb)	2033	4.150	0.7000	4.969	1.752	0.6259	31	-0.47	601.23	0
601	Aflatoxin B1 (ppb)	2369	4.170	0.4000	4.969	1.752	0.6259	31	-0.46	601.20	0
601	Aflatoxin B1 (ppb)	2369	4.170	0.4000	4.969	1.752	0.6259	31	-0.46	601.24	0
601	Aflatoxin B1 (ppb)	0028	4.360	0.3600	4.969	1.752	0.6259	31	-0.35	601.23	0
601	Aflatoxin B1 (ppb)	0047	4.390	2.040	4.969	1.752	0.6259	31	-0.33	601.16	0
601	Aflatoxin B1 (ppb)	0964	4.468	0.0039	4.969	1.752	0.6259	31	-0.29	601.16	0
601	Aflatoxin B1 (ppb)	0629	4.490	0.5600	4.969	1.752	0.6259	31	-0.27	601.20	0
601	Aflatoxin B1 (ppb)	0563	4.575	0.2700	4.969	1.752	0.6259	31	-0.22	601.20	0
601	Aflatoxin B1 (ppb)	0682	4.600	0.0000	4.969	1.752	0.6259	31	-0.21	601.23	0
601	Aflatoxin B1 (ppb)	2352	4.640	0.1800	4.969	1.752	0.6259	31	-0.19	601.23	0
601	Aflatoxin B1 (ppb)	0010	4.650	1.300	4.969	1.752	0.6259	31	-0.18	601.20	0
601	Aflatoxin B1 (ppb)	0860	4.653	0.0000	4.969	1.752	0.6259	31	-0.18	601.23	0
601	Aflatoxin B1 (ppb)	0505	4.875	0.7700	4.969	1.752	0.6259	31	-0.05	601.16	0
601	Aflatoxin B1 (ppb)	2192	5.400	0.2000	4.969	1.752	0.6259	31	0.25	601.20	0
601	Aflatoxin B1 (ppb)	2327	5.500	0.3600	4.969	1.752	0.6259	31	0.30	601.23	0
601	Aflatoxin B1 (ppb)	0009	5.690	2.200	4.969	1.752	0.6259	31	0.41	601.23	0
601	Aflatoxin B1 (ppb)	2411	5.745	0.9100	4.969	1.752	0.6259	31	0.44	601.23	0
601	Aflatoxin B1 (ppb)	0918	5.850	1.300	4.969	1.752	0.6259	31	0.50	601.23	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
601	Aflatoxin B1 (ppb)	0098	5.880	1.060	4.969	1.752	0.6259	31	0.52	601.24	0
601	Aflatoxin B1 (ppb)	2022	6.145	0.3100	4.969	1.752	0.6259	31	0.67	601.92	0
601	Aflatoxin B1 (ppb)	0218	6.350	0.3000	4.969	1.752	0.6259	31	0.79	601.23	0
601	Aflatoxin B1 (ppb)	2269	6.840	0.3600	4.969	1.752	0.6259	31	1.07	601.98	0
601	Aflatoxin B1 (ppb)	0043	7.127	2.919	4.969	1.752	0.6259	31	1.23	601.23	0
601	Aflatoxin B1 (ppb)	2022	7.735	0.0100	4.969	1.752	0.6259	31	1.58	601.91	0
601	Aflatoxin B1 (ppb)	2140	18.00	0.0000	4.969	1.752	0.6259	31	7.44	601.23	0
601	Aflatoxin B1 (ppb)	0242	21.50	9.000	4.969	1.752	0.6259	31	9.44	601.23	1
601	Aflatoxin B1 (ppb)	0032	0.0000	0.0000	4.969	1.752	0.6259	31		601.23	4
601	Aflatoxin B1 (ppb)	0553	< 1		4.969	1.752	0.6259	31		601.23	5
602	Aflatoxin B2 (ppb)	2246	0.1800	0.0000	0.3179	0.1207	0.0285	11	-1.14	602.24	0
602	Aflatoxin B2 (ppb)	0010	0.2000	0.0000	0.3179	0.1207	0.0285	11	-0.98	602.20	0
602	Aflatoxin B2 (ppb)	2411	0.2290	0.0020	0.3179	0.1207	0.0285	11	-0.74	602.23	0
602	Aflatoxin B2 (ppb)	0353	0.2925	0.0010	0.3179	0.1207	0.0285	11	-0.21	602.23	0
602	Aflatoxin B2 (ppb)	0629	0.3050	0.0100	0.3179	0.1207	0.0285	11	-0.11	602.20	0
602	Aflatoxin B2 (ppb)	0723	0.3350	0.0500	0.3179	0.1207	0.0285	11	0.14	602.20	0
602	Aflatoxin B2 (ppb)	0913	0.3500	0.1000	0.3179	0.1207	0.0285	11	0.27	602.23	0
602	Aflatoxin B2 (ppb)	2192	0.3850	0.0700	0.3179	0.1207	0.0285	11	0.56	602.20	0
602	Aflatoxin B2 (ppb)	0682	0.4000	0.0000	0.3179	0.1207	0.0285	11	0.68	602.23	0
602	Aflatoxin B2 (ppb)	2369	0.4100	0.0400	0.3179	0.1207	0.0285	11	0.76	602.20	0
602	Aflatoxin B2 (ppb)	2369	0.4100	0.0400	0.3179	0.1207	0.0285	11	0.76	602.24	0
602	Aflatoxin B2 (ppb)	0032	0.0000	0.0000	0.3179	0.1207	0.0285	11		602.23	4
602	Aflatoxin B2 (ppb)	2326	0.0000	0.0000	0.3179	0.1207	0.0285	11		602.23	4
602	Aflatoxin B2 (ppb)	2352	0.0000	0.0000	0.3179	0.1207	0.0285	11		602.23	4
602	Aflatoxin B2 (ppb)	0043	< 0.0279		0.3179	0.1207	0.0285	11		602.23	5
602	Aflatoxin B2 (ppb)	0028	< 0.295		0.3179	0.1207	0.0285	11		602.23	5
602	Aflatoxin B2 (ppb)	0870	< 0.6		0.3179	0.1207	0.0285	11		602.23	5
602	Aflatoxin B2 (ppb)	0218	< 1		0.3179	0.1207	0.0285	11		602.23	5
602	Aflatoxin B2 (ppb)	0242	< 1		0.3179	0.1207	0.0285	11		602.23	5
602	Aflatoxin B2 (ppb)	0553	< 1		0.3179	0.1207	0.0285	11		602.23	5
602	Aflatoxin B2 (ppb)	0918	< 1		0.3179	0.1207	0.0285	11		602.23	5
602	Aflatoxin B2 (ppb)	2033	< 1		0.3179	0.1207	0.0285	11		602.23	5
602	Aflatoxin B2 (ppb)	0098	< 1		0.3179	0.1207	0.0285	11		602.24	5
602	Aflatoxin B2 (ppb)	2327	< 1.2		0.3179	0.1207	0.0285	11		602.23	5
602	Aflatoxin B2 (ppb)	0563	< 5		0.3179	0.1207	0.0285	11		602.20	5
602	Aflatoxin B2 (ppb)	2140	< 20		0.3179	0.1207	0.0285	11		602.23	5
603	Aflatoxin G1 (ppb)	0964	0.6240	0.0266				2		603.16	0
603	Aflatoxin G1 (ppb)	0242	21.00	2.000				2		603.23	0
603	Aflatoxin G1 (ppb)	0032	0.0000	0.0000				2		603.23	4
603	Aflatoxin G1 (ppb)	0913	0.0000	0.0000				2		603.23	4
603	Aflatoxin G1 (ppb)	2326	0.0000	0.0000				2		603.23	4
603	Aflatoxin G1 (ppb)	2352	0.0000	0.0000				2		603.23	4

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
603	Aflatoxin G1 (ppb)	2411	< 0.01					2	603.23	5	
603	Aflatoxin G1 (ppb)	0043	< 0.1062					2	603.23	5	
603	Aflatoxin G1 (ppb)	0028	< 0.176					2	603.23	5	
603	Aflatoxin G1 (ppb)	2246	< 0.23					2	603.24	5	
603	Aflatoxin G1 (ppb)	0629	< 0.25					2	603.20	5	
603	Aflatoxin G1 (ppb)	0682	< 0.5					2	603.23	5	
603	Aflatoxin G1 (ppb)	0870	< 0.6					2	603.23	5	
603	Aflatoxin G1 (ppb)	0160	< 1					2	603.23	5	
603	Aflatoxin G1 (ppb)	0218	< 1					2	603.23	5	
603	Aflatoxin G1 (ppb)	0553	< 1					2	603.23	5	
603	Aflatoxin G1 (ppb)	0918	< 1					2	603.23	5	
603	Aflatoxin G1 (ppb)	2033	< 1					2	603.23	5	
603	Aflatoxin G1 (ppb)	2327	< 1.1					2	603.23	5	
603	Aflatoxin G1 (ppb)	0098	< 15					2	603.24	5	
603	Aflatoxin G1 (ppb)	2140	< 20					2	603.23	5	
604	Aflatoxin G2 (ppb)	2369	0.0650	0.0100				3	604.20	0	
604	Aflatoxin G2 (ppb)	2369	0.0650	0.0100				3	604.24	0	
604	Aflatoxin G2 (ppb)	2389	11.50	0.4000				3	604.98	0	
604	Aflatoxin G2 (ppb)	0032	0.0000	0.0000				3	604.23	4	
604	Aflatoxin G2 (ppb)	0913	0.0000	0.0000				3	604.23	4	
604	Aflatoxin G2 (ppb)	2326	0.0000	0.0000				3	604.23	4	
604	Aflatoxin G2 (ppb)	0910	< 0.01					3	604.21	5	
604	Aflatoxin G2 (ppb)	2246	< 0.01					3	604.24	5	
604	Aflatoxin G2 (ppb)	2411	< 0.17					3	604.23	5	
604	Aflatoxin G2 (ppb)	0629	< 0.25					3	604.20	5	
604	Aflatoxin G2 (ppb)	0028	< 0.332					3	604.23	5	
604	Aflatoxin G2 (ppb)	0682	< 0.5					3	604.23	5	
604	Aflatoxin G2 (ppb)	0160	< 1					3	604.23	5	
604	Aflatoxin G2 (ppb)	0218	< 1					3	604.23	5	
604	Aflatoxin G2 (ppb)	0242	< 1					3	604.23	5	
604	Aflatoxin G2 (ppb)	0553	< 1					3	604.23	5	
604	Aflatoxin G2 (ppb)	0870	< 1					3	604.23	5	
604	Aflatoxin G2 (ppb)	0918	< 1					3	604.23	5	
604	Aflatoxin G2 (ppb)	2033	< 1					3	604.23	5	
604	Aflatoxin G2 (ppb)	2327	< 1.6					3	604.23	5	
604	Aflatoxin G2 (ppb)	0098	< 5					3	604.24	5	
604	Aflatoxin G2 (ppb)	2140	< 20					3	604.23	5	
610	Deoxynivalenol (ppb)	2389	1.600	0.2000	1,185	360.1	101.9	57	-3.29	610.98	0
610	Deoxynivalenol (ppb)	0043	296.9	13.40	1,185	360.1	101.9	57	-2.47	610.23	0
610	Deoxynivalenol (ppb)	0553	562.8	6.900	1,185	360.1	101.9	57	-1.73	610.23	0
610	Deoxynivalenol (ppb)	0032	653.0	16.00	1,185	360.1	101.9	57	-1.48	610.23	0
610	Deoxynivalenol (ppb)	2326	744.0	66.00	1,185	360.1	101.9	57	-1.22	610.23	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
610	Deoxynivalenol (ppb)	0098	800.0	200.0	1,185	360.1	101.9	57	-1.07	610.20	0
610	Deoxynivalenol (ppb)	0913	800.0	0.0000	1,185	360.1	101.9	57	-1.07	610.23	0
610	Deoxynivalenol (ppb)	0066	890.0	180.0	1,185	360.1	101.9	57	-0.82	610.08	0
610	Deoxynivalenol (ppb)	2268	899.3	0.2810	1,185	360.1	101.9	57	-0.79	610.22	0
610	Deoxynivalenol (ppb)	0015	920.3	74.70	1,185	360.1	101.9	57	-0.73	610.16	0
610	Deoxynivalenol (ppb)	2129	923.2	8.700	1,185	360.1	101.9	57	-0.73	610.23	0
610	Deoxynivalenol (ppb)	2327	934.5	17.00	1,185	360.1	101.9	57	-0.69	610.23	0
610	Deoxynivalenol (ppb)	2355	935.0	50.00	1,185	360.1	101.9	57	-0.69	610.98	0
610	Deoxynivalenol (ppb)	2055	950.0	100.0	1,185	360.1	101.9	57	-0.65	610.23	0
610	Deoxynivalenol (ppb)	0870	961.3	25.00	1,185	360.1	101.9	57	-0.62	610.99	0
610	Deoxynivalenol (ppb)	0028	964.7	62.81	1,185	360.1	101.9	57	-0.61	610.23	0
610	Deoxynivalenol (ppb)	0218	981.5	19.00	1,185	360.1	101.9	57	-0.56	610.23	0
610	Deoxynivalenol (ppb)	2033	1,000	45.30	1,185	360.1	101.9	57	-0.51	610.23	0
610	Deoxynivalenol (ppb)	0009	1,006	68.21	1,185	360.1	101.9	57	-0.50	610.07	0
610	Deoxynivalenol (ppb)	2140	1,015	136.0	1,185	360.1	101.9	57	-0.47	610.23	0
610	Deoxynivalenol (ppb)	2352	1,019	36.97	1,185	360.1	101.9	57	-0.46	610.23	0
610	Deoxynivalenol (ppb)	0596	1,030	140.0	1,185	360.1	101.9	57	-0.43	610.98	0
610	Deoxynivalenol (ppb)	0413	1,050	100.0	1,185	360.1	101.9	57	-0.37	610.01	0
610	Deoxynivalenol (ppb)	0010	1,051	57.00	1,185	360.1	101.9	57	-0.37	610.23	0
610	Deoxynivalenol (ppb)	0918	1,071	63.14	1,185	360.1	101.9	57	-0.32	610.23	0
610	Deoxynivalenol (ppb)	0353	1,080	2.000	1,185	360.1	101.9	57	-0.29	610.23	0
610	Deoxynivalenol (ppb)	2269	1,093	35.00	1,185	360.1	101.9	57	-0.26	610.98	0
610	Deoxynivalenol (ppb)	2098	1,120	100.0	1,185	360.1	101.9	57	-0.18	610.01	0
610	Deoxynivalenol (ppb)	0682	1,130	0.0000	1,185	360.1	101.9	57	-0.15	610.23	0
610	Deoxynivalenol (ppb)	0870	1,140	40.00	1,185	360.1	101.9	57	-0.12	610.23	0
610	Deoxynivalenol (ppb)	0511	1,150	300.0	1,185	360.1	101.9	57	-0.10	610.01	0
610	Deoxynivalenol (ppb)	0413	1,150	100.0	1,185	360.1	101.9	57	-0.10	610.20	0
610	Deoxynivalenol (ppb)	2411	1,150	52.00	1,185	360.1	101.9	57	-0.10	610.23	0
610	Deoxynivalenol (ppb)	0918	1,219	50.00	1,185	360.1	101.9	57	0.10	610.01	0
610	Deoxynivalenol (ppb)	0208	1,240	100.0	1,185	360.1	101.9	57	0.15	610.01	0
610	Deoxynivalenol (ppb)	0529	1,250	100.0	1,185	360.1	101.9	57	0.18	610.01	0
610	Deoxynivalenol (ppb)	2113	1,250	300.0	1,185	360.1	101.9	57	0.18	610.01	0
610	Deoxynivalenol (ppb)	0297	1,300	0.0000	1,185	360.1	101.9	57	0.32	610.01	0
610	Deoxynivalenol (ppb)	2192	1,330	100.0	1,185	360.1	101.9	57	0.40	610.01	0
610	Deoxynivalenol (ppb)	0964	1,343	39.47	1,185	360.1	101.9	57	0.44	610.16	0
610	Deoxynivalenol (ppb)	0014	1,350	100.0	1,185	360.1	101.9	57	0.46	610.01	0
610	Deoxynivalenol (ppb)	0042	1,350	100.0	1,185	360.1	101.9	57	0.46	610.01	0
610	Deoxynivalenol (ppb)	0407	1,396	231.0	1,185	360.1	101.9	57	0.59	610.07	0
610	Deoxynivalenol (ppb)	0590	1,470	320.0	1,185	360.1	101.9	57	0.79	610.07	0
610	Deoxynivalenol (ppb)	0505	1,476	305.6	1,185	360.1	101.9	57	0.81	610.16	0
610	Deoxynivalenol (ppb)	0033	1,503	255.0	1,185	360.1	101.9	57	0.88	610.01	0
610	Deoxynivalenol (ppb)	0027	1,585	490.0	1,185	360.1	101.9	57	1.11	610.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
610	Deoxynivalenol (ppb)	0227	1,600	200.0	1,185	360.1	101.9	57	1.15	610.01	0
610	Deoxynivalenol (ppb)	0242	1,612	379.0	1,185	360.1	101.9	57	1.19	610.23	0
610	Deoxynivalenol (ppb)	0723	1,637	17.30	1,185	360.1	101.9	57	1.26	610.05	0
610	Deoxynivalenol (ppb)	2022	1,715	70.00	1,185	360.1	101.9	57	1.47	610.01	0
610	Deoxynivalenol (ppb)	2022	1,730	80.00	1,185	360.1	101.9	57	1.51	610.05	0
610	Deoxynivalenol (ppb)	2369	1,736	19.12	1,185	360.1	101.9	57	1.53	610.09	0
610	Deoxynivalenol (ppb)	2369	1,736	19.12	1,185	360.1	101.9	57	1.53	610.20	0
610	Deoxynivalenol (ppb)	0142	1,750	100.0	1,185	360.1	101.9	57	1.57	610.98	0
610	Deoxynivalenol (ppb)	0563	1,924	302.0	1,185	360.1	101.9	57	2.05	610.25	0
610	Deoxynivalenol (ppb)	0035	2,035	14.00	1,185	360.1	101.9	57	2.36	610.08	0
610	Deoxynivalenol (ppb)	2411	5,645	1,770	1,185	360.1	101.9	57	12.39	610.20	1
610	Deoxynivalenol (ppb)	2246	24,200	1,000	1,185	360.1	101.9	57	63.91	610.05	2
620	Total Fumonisin (ppb)	2411	516.0	0.0000	1,174	356.9	130.8	36	-1.84	620.20	0
620	Total Fumonisin (ppb)	0042	750.0	300.0	1,174	356.9	130.8	36	-1.19	620.01	0
620	Total Fumonisin (ppb)	0098	750.0	100.0	1,174	356.9	130.8	36	-1.19	620.11	0
620	Total Fumonisin (ppb)	0407	802.0	98.00	1,174	356.9	130.8	36	-1.04	620.09	0
620	Total Fumonisin (ppb)	0010	805.0	170.0	1,174	356.9	130.8	36	-1.03	620.23	0
620	Total Fumonisin (ppb)	0918	818.0	176.0	1,174	356.9	130.8	36	-1.00	620.01	0
620	Total Fumonisin (ppb)	0027	890.0	600.0	1,174	356.9	130.8	36	-0.80	620.01	0
620	Total Fumonisin (ppb)	0520	905.0	50.00	1,174	356.9	130.8	36	-0.75	620.12	0
620	Total Fumonisin (ppb)	2369	981.5	72.00	1,174	356.9	130.8	36	-0.54	620.11	0
620	Total Fumonisin (ppb)	0682	991.7	0.0000	1,174	356.9	130.8	36	-0.51	620.23	0
620	Total Fumonisin (ppb)	0511	1,000	200.0	1,174	356.9	130.8	36	-0.49	620.01	0
620	Total Fumonisin (ppb)	2022	1,010	0.0000	1,174	356.9	130.8	36	-0.46	620.07	0
620	Total Fumonisin (ppb)	2033	1,018	142.3	1,174	356.9	130.8	36	-0.44	620.23	0
620	Total Fumonisin (ppb)	2352	1,018	0.7500	1,174	356.9	130.8	36	-0.44	620.23	0
620	Total Fumonisin (ppb)	2113	1,050	100.0	1,174	356.9	130.8	36	-0.35	620.01	0
620	Total Fumonisin (ppb)	2098	1,090	40.00	1,174	356.9	130.8	36	-0.24	620.02	0
620	Total Fumonisin (ppb)	0014	1,095	110.0	1,174	356.9	130.8	36	-0.22	620.01	0
620	Total Fumonisin (ppb)	0015	1,141	71.20	1,174	356.9	130.8	36	-0.09	620.16	0
620	Total Fumonisin (ppb)	2022	1,145	70.00	1,174	356.9	130.8	36	-0.08	620.02	0
620	Total Fumonisin (ppb)	0964	1,172	109.8	1,174	356.9	130.8	36	0.00	620.16	0
620	Total Fumonisin (ppb)	0723	1,230	20.00	1,174	356.9	130.8	36	0.16	620.07	0
620	Total Fumonisin (ppb)	0035	1,252	5.000	1,174	356.9	130.8	36	0.22	620.09	0
620	Total Fumonisin (ppb)	0505	1,261	35.65	1,174	356.9	130.8	36	0.24	620.16	0
620	Total Fumonisin (ppb)	2411	1,287	104.9	1,174	356.9	130.8	36	0.32	620.23	0
620	Total Fumonisin (ppb)	2192	1,295	70.00	1,174	356.9	130.8	36	0.34	620.01	0
620	Total Fumonisin (ppb)	0033	1,335	173.0	1,174	356.9	130.8	36	0.45	620.01	0
620	Total Fumonisin (ppb)	0032	1,415	10.00	1,174	356.9	130.8	36	0.68	620.23	0
620	Total Fumonisin (ppb)	0353	1,415	64.00	1,174	356.9	130.8	36	0.68	620.23	0
620	Total Fumonisin (ppb)	2268	1,450	500.0	1,174	356.9	130.8	36	0.77	620.11	0
620	Total Fumonisin (ppb)	0066	1,490	20.00	1,174	356.9	130.8	36	0.89	620.09	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
620	Total Fumonisin (ppb)	0297	1,500	0.0000	1,174	356.9	130.8	36	0.91	620.01	0
620	Total Fumonisin (ppb)	0529	1,550	300.0	1,174	356.9	130.8	36	1.05	620.02	0
620	Total Fumonisin (ppb)	0596	2,400	600.0	1,174	356.9	130.8	36	3.44	620.98	0
620	Total Fumonisin (ppb)	2129	2,472	96.00	1,174	356.9	130.8	36	3.64	620.23	0
620	Total Fumonisin (ppb)	0208	2,550	100.0	1,174	356.9	130.8	36	3.86	620.98	0
620	Total Fumonisin (ppb)	0142	2,600	200.0	1,174	356.9	130.8	36	4.00	620.98	0
620	Total Fumonisin (ppb)	2246	10,330	1,840	1,174	356.9	130.8	36	25.65	620.07	2
620	Total Fumonisin (ppb)	0227	< 1000		1,174	356.9	130.8	36		620.01	5
621	Fumonsin B1 (ppb)	0043	105.4	0.5000	700.5	216.0	91.47	23	-2.76	621.23	0
621	Fumonsin B1 (ppb)	0009	218.3	134.3	700.5	216.0	91.47	23	-2.23	621.23	0
621	Fumonsin B1 (ppb)	2411	389.0	0.0000	700.5	216.0	91.47	23	-1.44	621.20	0
621	Fumonsin B1 (ppb)	0913	450.0	100.0	700.5	216.0	91.47	23	-1.16	621.23	0
621	Fumonsin B1 (ppb)	2352	527.6	2.810	700.5	216.0	91.47	23	-0.80	621.23	0
621	Fumonsin B1 (ppb)	0098	550.0	100.0	700.5	216.0	91.47	23	-0.70	621.21	0
621	Fumonsin B1 (ppb)	2140	569.5	37.00	700.5	216.0	91.47	23	-0.61	621.23	0
621	Fumonsin B1 (ppb)	2327	587.5	13.00	700.5	216.0	91.47	23	-0.52	621.23	0
621	Fumonsin B1 (ppb)	0010	590.0	120.0	700.5	216.0	91.47	23	-0.51	621.23	0
621	Fumonsin B1 (ppb)	2326	623.0	214.0	700.5	216.0	91.47	23	-0.36	621.23	0
621	Fumonsin B1 (ppb)	0218	661.0	24.00	700.5	216.0	91.47	23	-0.18	621.23	0
621	Fumonsin B1 (ppb)	0918	699.0	18.18	700.5	216.0	91.47	23	-0.01	621.23	0
621	Fumonsin B1 (ppb)	2033	703.3	80.60	700.5	216.0	91.47	23	0.01	621.23	0
621	Fumonsin B1 (ppb)	0682	710.0	0.0000	700.5	216.0	91.47	23	0.04	621.23	0
621	Fumonsin B1 (ppb)	0870	730.0	20.00	700.5	216.0	91.47	23	0.14	621.23	0
621	Fumonsin B1 (ppb)	2369	740.8	135.4	700.5	216.0	91.47	23	0.19	621.21	0
621	Fumonsin B1 (ppb)	2055	900.0	200.0	700.5	216.0	91.47	23	0.92	621.23	0
621	Fumonsin B1 (ppb)	0353	918.5	165.0	700.5	216.0	91.47	23	1.01	621.23	0
621	Fumonsin B1 (ppb)	2411	934.5	113.0	700.5	216.0	91.47	23	1.08	621.23	0
621	Fumonsin B1 (ppb)	0032	1,005	10.00	700.5	216.0	91.47	23	1.41	621.23	0
621	Fumonsin B1 (ppb)	0563	1,020	110.0	700.5	216.0	91.47	23	1.48	621.20	0
621	Fumonsin B1 (ppb)	0553	1,226	206.0	700.5	216.0	91.47	23	2.43	621.23	0
621	Fumonsin B1 (ppb)	2355	2,850	300.0	700.5	216.0	91.47	23	9.95	621.98	0
622	Fumonisin B2 (ppb)	0913	100.0	0.0000	248.3	78.73	29.63	18	-1.88	622.23	0
622	Fumonisin B2 (ppb)	2411	127.0	0.0000	248.3	78.73	29.63	18	-1.54	622.20	0
622	Fumonisin B2 (ppb)	0010	165.0	30.00	248.3	78.73	29.63	18	-1.06	622.23	0
622	Fumonisin B2 (ppb)	0098	200.0	0.0000	248.3	78.73	29.63	18	-0.61	622.21	0
622	Fumonisin B2 (ppb)	2326	208.0	72.00	248.3	78.73	29.63	18	-0.51	622.23	0
622	Fumonisin B2 (ppb)	0218	234.5	35.00	248.3	78.73	29.63	18	-0.17	622.23	0
622	Fumonisin B2 (ppb)	2369	240.7	63.40	248.3	78.73	29.63	18	-0.10	622.21	0
622	Fumonisin B2 (ppb)	2140	241.0	18.00	248.3	78.73	29.63	18	-0.09	622.23	0
622	Fumonisin B2 (ppb)	2033	249.5	53.00	248.3	78.73	29.63	18	0.02	622.23	0
622	Fumonisin B2 (ppb)	0870	250.0	20.00	248.3	78.73	29.63	18	0.02	622.23	0
622	Fumonisin B2 (ppb)	2055	250.0	100.0	248.3	78.73	29.63	18	0.02	622.23	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
622	Fumonisin B2 (ppb)	2327	253.5	21.00	248.3	78.73	29.63	18	0.07	622.23	0
622	Fumonisin B2 (ppb)	0682	280.0	0.0000	248.3	78.73	29.63	18	0.40	622.23	0
622	Fumonisin B2 (ppb)	2411	287.0	0.0000	248.3	78.73	29.63	18	0.49	622.23	0
622	Fumonisin B2 (ppb)	0353	289.0	24.00	248.3	78.73	29.63	18	0.52	622.23	0
622	Fumonisin B2 (ppb)	0032	336.0	12.00	248.3	78.73	29.63	18	1.11	622.23	0
622	Fumonisin B2 (ppb)	0918	346.7	37.08	248.3	78.73	29.63	18	1.25	622.23	0
622	Fumonisin B2 (ppb)	0553	410.5	47.80	248.3	78.73	29.63	18	2.06	622.23	0
622	Fumonisin B2 (ppb)	0043	< 3.4088		248.3	78.73	29.63	18		622.23	5
623	Fumonisin B3 (ppb)	0010	50.00	20.00	78.01	25.53	15.59	9	-1.10	623.23	0
623	Fumonisin B3 (ppb)	2352	56.81	0.6500	78.01	25.53	15.59	9	-0.83	623.23	0
623	Fumonisin B3 (ppb)	2033	64.85	8.700	78.01	25.53	15.59	9	-0.52	623.23	0
623	Fumonisin B3 (ppb)	2411	65.35	8.100	78.01	25.53	15.59	9	-0.50	623.23	0
623	Fumonisin B3 (ppb)	0032	69.50	2.200	78.01	25.53	15.59	9	-0.33	623.23	0
623	Fumonisin B3 (ppb)	2327	71.20	2.400	78.01	25.53	15.59	9	-0.27	623.23	0
623	Fumonisin B3 (ppb)	0353	74.00	8.000	78.01	25.53	15.59	9	-0.16	623.23	0
623	Fumonisin B3 (ppb)	2326	133.0	32.00	78.01	25.53	15.59	9	2.15	623.23	0
623	Fumonisin B3 (ppb)	0553	265.3	58.30	78.01	25.53	15.59	9	7.34	623.23	0
623	Fumonisin B3 (ppb)	0913	0.0000	0.0000	78.01	25.53	15.59	9		623.23	4
623	Fumonisin B3 (ppb)	0870	< 100		78.01	25.53	15.59	9		623.23	5
630	Ochratoxin A (ppb)	0596	3.900	0.2000	13.37	4.589	4.224	38	-2.06	630.98	0
630	Ochratoxin A (ppb)	2098	4.450	0.1000	13.37	4.589	4.224	38	-1.94	630.01	0
630	Ochratoxin A (ppb)	2226	4.767	0.4120	13.37	4.589	4.224	38	-1.87	630.05	0
630	Ochratoxin A (ppb)	0870	6.430	0.5400	13.37	4.589	4.224	38	-1.51	630.23	0
630	Ochratoxin A (ppb)	0407	6.710	1.220	13.37	4.589	4.224	38	-1.45	630.04	0
630	Ochratoxin A (ppb)	0629	7.445	0.3500	13.37	4.589	4.224	38	-1.29	630.20	0
630	Ochratoxin A (ppb)	2033	8.200	0.8000	13.37	4.589	4.224	38	-1.13	630.23	0
630	Ochratoxin A (ppb)	0918	8.250	0.7000	13.37	4.589	4.224	38	-1.12	630.23	0
630	Ochratoxin A (ppb)	0047	8.530	4.680	13.37	4.589	4.224	38	-1.05	630.16	0
630	Ochratoxin A (ppb)	0723	10.38	0.4100	13.37	4.589	4.224	38	-0.65	630.03	0
630	Ochratoxin A (ppb)	0015	10.55	1.100	13.37	4.589	4.224	38	-0.61	630.16	0
630	Ochratoxin A (ppb)	0010	10.85	0.7000	13.37	4.589	4.224	38	-0.55	630.23	0
630	Ochratoxin A (ppb)	0218	11.50	1.000	13.37	4.589	4.224	38	-0.41	630.23	0
630	Ochratoxin A (ppb)	2369	11.55	0.4200	13.37	4.589	4.224	38	-0.40	630.06	0
630	Ochratoxin A (ppb)	2369	11.55	0.4200	13.37	4.589	4.224	38	-0.40	630.20	0
630	Ochratoxin A (ppb)	0913	11.55	2.300	13.37	4.589	4.224	38	-0.40	630.23	0
630	Ochratoxin A (ppb)	0860	11.84	0.0000	13.37	4.589	4.224	38	-0.33	630.23	0
630	Ochratoxin A (ppb)	0353	12.50	3.000	13.37	4.589	4.224	38	-0.19	630.23	0
630	Ochratoxin A (ppb)	2022	13.20	0.0600	13.37	4.589	4.224	38	-0.04	630.01	0
630	Ochratoxin A (ppb)	0098	13.50	1.400	13.37	4.589	4.224	38	0.03	630.20	0
630	Ochratoxin A (ppb)	0682	13.60	0.0000	13.37	4.589	4.224	38	0.05	630.23	0
630	Ochratoxin A (ppb)	0563	13.74	0.1900	13.37	4.589	4.224	38	0.08	630.20	0
630	Ochratoxin A (ppb)	0505	13.85	0.5700	13.37	4.589	4.224	38	0.10	630.16	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
630	Ochratoxin A (ppb)	2022	13.99	1.600	13.37	4.589	4.224	38	0.14	630.03	0
630	Ochratoxin A (ppb)	0520	14.00	0.0000	13.37	4.589	4.224	38	0.14	630.06	0
630	Ochratoxin A (ppb)	0066	14.43	0.6500	13.37	4.589	4.224	38	0.23	630.05	0
630	Ochratoxin A (ppb)	0032	15.70	3.200	13.37	4.589	4.224	38	0.51	630.23	0
630	Ochratoxin A (ppb)	2326	16.00	2.000	13.37	4.589	4.224	38	0.57	630.23	0
630	Ochratoxin A (ppb)	0964	16.99	3.050	13.37	4.589	4.224	38	0.79	630.16	0
630	Ochratoxin A (ppb)	2327	17.75	0.1000	13.37	4.589	4.224	38	0.95	630.23	0
630	Ochratoxin A (ppb)	2129	18.45	1.700	13.37	4.589	4.224	38	1.11	630.23	0
630	Ochratoxin A (ppb)	2352	19.67	0.1000	13.37	4.589	4.224	38	1.37	630.23	0
630	Ochratoxin A (ppb)	2411	19.85	7.100	13.37	4.589	4.224	38	1.41	630.23	0
630	Ochratoxin A (ppb)	2192	29.00	0.0000	13.37	4.589	4.224	38	3.41	630.20	0
630	Ochratoxin A (ppb)	2246	29.14	1.950	13.37	4.589	4.224	38	3.44	630.03	0
630	Ochratoxin A (ppb)	2055	29.45	5.300	13.37	4.589	4.224	38	3.50	630.23	0
630	Ochratoxin A (ppb)	0529	50.75	8.500	13.37	4.589	4.224	38	8.15	630.01	0
630	Ochratoxin A (ppb)	0723	151.1	104.7	13.37	4.589	4.224	38	30.01	630.20	0
630	Ochratoxin A (ppb)	0918	567.0	140.0	13.37	4.589	4.224	38	120.65	630.01	2
630	Ochratoxin A (ppb)	0553	< 1		13.37	4.589	4.224	38		630.23	5
630	Ochratoxin A (ppb)	2140	< 20		13.37	4.589	4.224	38		630.23	5
640	T-2 Toxin (ppb)	2352	68.23	0.6000	120.3	38.89	10.53	21	-1.34	640.23	0
640	T-2 Toxin (ppb)	2326	73.50	1.000	120.3	38.89	10.53	21	-1.20	640.23	0
640	T-2 Toxin (ppb)	2033	81.25	10.70	120.3	38.89	10.53	21	-1.00	640.23	0
640	T-2 Toxin (ppb)	2411	85.70	3.200	120.3	38.89	10.53	21	-0.89	640.23	0
640	T-2 Toxin (ppb)	0870	90.91	8.940	120.3	38.89	10.53	21	-0.75	640.23	0
640	T-2 Toxin (ppb)	2055	91.50	2.000	120.3	38.89	10.53	21	-0.74	640.23	0
640	T-2 Toxin (ppb)	0913	92.00	3.600	120.3	38.89	10.53	21	-0.73	640.23	0
640	T-2 Toxin (ppb)	2327	94.55	3.700	120.3	38.89	10.53	21	-0.66	640.23	0
640	T-2 Toxin (ppb)	0918	97.38	6.420	120.3	38.89	10.53	21	-0.59	640.23	0
640	T-2 Toxin (ppb)	0353	100.5	1.000	120.3	38.89	10.53	21	-0.51	640.23	0
640	T-2 Toxin (ppb)	0218	102.0	4.000	120.3	38.89	10.53	21	-0.47	640.23	0
640	T-2 Toxin (ppb)	0015	107.5	31.70	120.3	38.89	10.53	21	-0.33	640.16	0
640	T-2 Toxin (ppb)	2140	118.5	69.00	120.3	38.89	10.53	21	-0.05	640.23	0
640	T-2 Toxin (ppb)	0407	121.5	3.160	120.3	38.89	10.53	21	0.03	640.05	0
640	T-2 Toxin (ppb)	0682	141.6	0.0000	120.3	38.89	10.53	21	0.55	640.23	0
640	T-2 Toxin (ppb)	2129	148.9	3.200	120.3	38.89	10.53	21	0.74	640.23	0
640	T-2 Toxin (ppb)	0297	167.7	3.600	120.3	38.89	10.53	21	1.22	640.98	0
640	T-2 Toxin (ppb)	2269	174.5	22.50	120.3	38.89	10.53	21	1.39	640.98	0
640	T-2 Toxin (ppb)	0066	192.4	5.870	120.3	38.89	10.53	21	1.86	640.03	0
640	T-2 Toxin (ppb)	0563	200.0	34.00	120.3	38.89	10.53	21	2.05	640.25	0
640	T-2 Toxin (ppb)	0918	222.3	3.000	120.3	38.89	10.53	21	2.62	640.98	0
640	T-2 Toxin (ppb)	2246	848.1	7.498	120.3	38.89	10.53	21	18.71	640.03	2
640	T-2 Toxin (ppb)	0553	< 1		120.3	38.89	10.53	21		640.23	5
641	HT-2 Toxin (ppb)	0870	34.89	3.940	44.80	14.88	5.542	9	-0.67	641.23	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
641	HT-2 Toxin (ppb)	0913	36.30	1.400	44.80	14.88	5.542	9	-0.57	641.23	0
641	HT-2 Toxin (ppb)	2411	38.90	8.000	44.80	14.88	5.542	9	-0.40	641.23	0
641	HT-2 Toxin (ppb)	2326	39.50	5.000	44.80	14.88	5.542	9	-0.36	641.23	0
641	HT-2 Toxin (ppb)	2033	39.55	0.7000	44.80	14.88	5.542	9	-0.35	641.23	0
641	HT-2 Toxin (ppb)	0218	40.50	7.000	44.80	14.88	5.542	9	-0.29	641.23	0
641	HT-2 Toxin (ppb)	2352	48.48	1.840	44.80	14.88	5.542	9	0.25	641.23	0
641	HT-2 Toxin (ppb)	0353	62.00	6.000	44.80	14.88	5.542	9	1.16	641.23	0
641	HT-2 Toxin (ppb)	2140	119.0	16.00	44.80	14.88	5.542	9	4.99	641.23	0
641	HT-2 Toxin (ppb)	2327	< 200		44.80	14.88	5.542	9		641.23	5
642	T-2 Toxin + HT-2 Toxin (ppb)	0142	112.1	3.000	167.1	53.57	14.61	13	-1.03	642.98	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0596	115.9	7.900	167.1	53.57	14.61	13	-0.96	642.98	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0590	126.0	34.00	167.1	53.57	14.61	13	-0.77	642.98	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0407	128.0	6.890	167.1	53.57	14.61	13	-0.73	642.99	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0227	138.0	8.000	167.1	53.57	14.61	13	-0.54	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0964	148.6	22.64	167.1	53.57	14.61	13	-0.35	642.16	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2098	159.2	3.200	167.1	53.57	14.61	13	-0.15	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0353	161.0	2.000	167.1	53.57	14.61	13	-0.11	642.04	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0529	164.1	10.00	167.1	53.57	14.61	13	-0.06	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2192	219.5	53.00	167.1	53.57	14.61	13	0.98	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2022	226.9	4.310	167.1	53.57	14.61	13	1.12	642.98	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2022	232.6	3.940	167.1	53.57	14.61	13	1.22	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2269	240.8	31.10	167.1	53.57	14.61	13	1.37	642.98	0
650	Zearalenone (ppb)	2389	28.30	0.2000	214.1	68.16	17.29	46	-2.73	650.98	0
650	Zearalenone (ppb)	2268	69.50	7.000	214.1	68.16	17.29	46	-2.12	650.99	0
650	Zearalenone (ppb)	0520	81.50	29.00	214.1	68.16	17.29	46	-1.94	650.99	0
650	Zearalenone (ppb)	0043	126.2	52.25	214.1	68.16	17.29	46	-1.29	650.24	0
650	Zearalenone (ppb)	0015	134.2	25.00	214.1	68.16	17.29	46	-1.17	650.16	0
650	Zearalenone (ppb)	2098	143.0	13.60	214.1	68.16	17.29	46	-1.04	650.01	0
650	Zearalenone (ppb)	0913	144.2	10.20	214.1	68.16	17.29	46	-1.03	650.24	0
650	Zearalenone (ppb)	0242	144.4	13.14	214.1	68.16	17.29	46	-1.02	650.24	0
650	Zearalenone (ppb)	0047	145.1	13.50	214.1	68.16	17.29	46	-1.01	650.16	0
650	Zearalenone (ppb)	2192	146.0	6.000	214.1	68.16	17.29	46	-1.00	650.20	0
650	Zearalenone (ppb)	0918	153.5	1.700	214.1	68.16	17.29	46	-0.89	650.01	0
650	Zearalenone (ppb)	0682	154.9	0.0000	214.1	68.16	17.29	46	-0.87	650.24	0
650	Zearalenone (ppb)	0032	155.5	15.00	214.1	68.16	17.29	46	-0.86	650.24	0
650	Zearalenone (ppb)	0870	160.1	5.453	214.1	68.16	17.29	46	-0.79	650.01	0
650	Zearalenone (ppb)	0918	170.5	2.400	214.1	68.16	17.29	46	-0.64	650.24	0
650	Zearalenone (ppb)	2033	174.1	2.100	214.1	68.16	17.29	46	-0.59	650.24	0
650	Zearalenone (ppb)	2129	175.0	4.000	214.1	68.16	17.29	46	-0.57	650.24	0
650	Zearalenone (ppb)	2140	178.5	5.000	214.1	68.16	17.29	46	-0.52	650.24	0
650	Zearalenone (ppb)	2369	179.2	14.47	214.1	68.16	17.29	46	-0.51	650.20	0
650	Zearalenone (ppb)	2369	179.2	14.47	214.1	68.16	17.29	46	-0.51	650.21	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Method Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob Mean	Rev.Horwitz	R-bar	# Tests			
650	Zearalenone (ppb)	0066	188.9	17.33	214.1	68.16	17.29	46	-0.37	650.07	0
650	Zearalenone (ppb)	0027	189.2	21.20	214.1	68.16	17.29	46	-0.36	650.01	0
650	Zearalenone (ppb)	0964	192.0	60.74	214.1	68.16	17.29	46	-0.32	650.16	0
650	Zearalenone (ppb)	2113	195.0	10.00	214.1	68.16	17.29	46	-0.28	650.01	0
650	Zearalenone (ppb)	0353	207.5	29.00	214.1	68.16	17.29	46	-0.10	650.23	0
650	Zearalenone (ppb)	0010	222.0	12.00	214.1	68.16	17.29	46	0.12	650.24	0
650	Zearalenone (ppb)	0098	230.0	20.00	214.1	68.16	17.29	46	0.23	650.20	0
650	Zearalenone (ppb)	0413	230.0	20.00	214.1	68.16	17.29	46	0.23	650.20	0
650	Zearalenone (ppb)	0563	232.5	5.000	214.1	68.16	17.29	46	0.27	650.20	0
650	Zearalenone (ppb)	0723	238.0	5.730	214.1	68.16	17.29	46	0.35	650.20	0
650	Zearalenone (ppb)	2022	238.6	21.88	214.1	68.16	17.29	46	0.36	650.01	0
650	Zearalenone (ppb)	0218	242.0	0.0000	214.1	68.16	17.29	46	0.41	650.24	0
650	Zearalenone (ppb)	2226	248.4	5.129	214.1	68.16	17.29	46	0.50	650.06	0
650	Zearalenone (ppb)	2269	258.2	24.50	214.1	68.16	17.29	46	0.65	650.98	0
650	Zearalenone (ppb)	0529	271.4	18.60	214.1	68.16	17.29	46	0.84	650.01	0
650	Zearalenone (ppb)	0297	275.6	10.10	214.1	68.16	17.29	46	0.90	650.01	0
650	Zearalenone (ppb)	2022	277.9	20.43	214.1	68.16	17.29	46	0.94	650.04	0
650	Zearalenone (ppb)	0590	293.5	9.000	214.1	68.16	17.29	46	1.17	650.06	0
650	Zearalenone (ppb)	0142	305.3	36.10	214.1	68.16	17.29	46	1.34	650.98	0
650	Zearalenone (ppb)	0035	334.2	69.70	214.1	68.16	17.29	46	1.76	650.07	0
650	Zearalenone (ppb)	2327	364.0	4.000	214.1	68.16	17.29	46	2.20	650.24	0
650	Zearalenone (ppb)	0596	365.8	71.60	214.1	68.16	17.29	46	2.23	650.98	0
650	Zearalenone (ppb)	2352	403.3	21.45	214.1	68.16	17.29	46	2.78	650.24	0
650	Zearalenone (ppb)	0407	472.1	8.910	214.1	68.16	17.29	46	3.79	650.07	0
650	Zearalenone (ppb)	2246	509.0	31.17	214.1	68.16	17.29	46	4.33	650.04	0
650	Zearalenone (ppb)	0723	885.5	7.160	214.1	68.16	17.29	46	9.85	650.04	0
650	Zearalenone (ppb)	0553	< 1		214.1	68.16	17.29	46		650.24	5

Notes: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2 and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = Rejected for duplicates too far apart, 2 = Rejected as outlier, 5 = A reporting limit and 4 = zeros submitted as values. Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). Flag 3 indicates not used in statistics.