



AAFCO
Association of American Feed Control Officials



Minerals Scheme

Rabbit Feed, Residue

Test Material Code # 202352

Analyte Summary Report

Labs Reporting: 30

Analytes Reported: 16

Issue Date : 07/31/2023

Analyte Group Code	Analyte and Method	Total # Tests Submitted	# Tests in Robust Calcs	Raw Mean	Raw SD	Assigned Value Robust Mean	Horwitz SD	Uncertainty (U) Robust	% RSD - Horwitz	Average Range (Rob R-bar)	Thompson Horwitz %RSD
015	Aluminum (ppm)	15	15	714.0	136.2	722.5	42.92	39.25	5.94%	24.44	5.94%
017	Boron (ppm)	7	7	59.18	4.003	59.76	5.166	1.456	8.64%	1.051	8.64%
021	Cobalt (ppm)	21	21	10.69	2.363	10.87	1.214	0.6197	11.17%	0.3750	11.17%
022	Copper (ppm)	26	25	134.7	29.76	136.0	10.39	7.673	7.64%	3.228	7.64%
023	Fluorine (ppm)	2	2	33.56	19.01						
024	Iodine (ppm)	2	2	85.95	19.76						
034	Selenium, Total (Se) (ppm)	18	14	9.456	21.23	1.228	0.1905	0.1663	15.51%	0.0825	15.51%
036	Sulfur (%)	11	11	0.3435	0.0290	0.3436	0.0161	0.0123	4.70%	0.0098	4.70%
038	Molybdenum (ppm)	20	20	47.90	5.114	47.91	4.281	1.577	8.94%	1.741	8.94%
041	Vanadium (ppm)	9	7	1.904	0.2196	1.904	0.2765	0.1177	14.52%	0.0669	14.52%
516	Arsenic, Total (As) (ppm)	21	18	3.728	1.031	3.709	0.4871	0.2020	13.13%	0.1066	13.13%
518	Cadmium (ppm)	22	19	0.6702	0.1908	0.6986	0.1179	0.0417	16.88%	0.0247	16.88%
520	Chromium, Total (Cr) (ppm)	22	22	52.21	21.45	54.82	4.800	4.788	8.76%	1.710	8.76%
526	Lead (ppm)	25	24	55.86	16.88	60.30	5.205	1.700	8.63%	1.171	8.63%
529	Mercury (ppb)	11	9	121.1	92.75	121.1	26.61	43.83	21.98%	5.423	21.98%
539	Nickel (ppm)	18	18	54.70	11.42	56.07	4.893	2.384	8.73%	2.387	8.73%

Notes: Robust statistics not used if < 6 labs reporting. In this case Means and SD's may be reported based on Raw Data with obvious blunders removed. Robust Assigned Values indicated in bold font. Horwitz and Revised Horwitz SD's assigned regardless of participant number.