



AAFCO
Association of American Feed Control Officials



**Mycotoxin Contaminants Scheme
Sheep Feed**

**# Labs Reporting: 64
Analytes Reported 15
Issue Date : 08/31/2023**

Test Material Code # 202363

Analyte Proficiency Testing Report

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests			
600	Total Aflatoxin (ppb)	2268	1.025	0.1500	3.897	1.383	0.4614	45	-2.08	600.14	0
600	Total Aflatoxin (ppb)	2325	1.380	0.0200	3.897	1.383	0.4614	45	-1.82	600.22	0
600	Total Aflatoxin (ppb)	2345	1.800	0.4000	3.897	1.383	0.4614	45	-1.52	600.14	0
600	Total Aflatoxin (ppb)	2322	2.060	0.3600	3.897	1.383	0.4614	45	-1.33	600.20	0
600	Total Aflatoxin (ppb)	0876	2.200	0.0000	3.897	1.383	0.4614	45	-1.23	600.01	0
600	Total Aflatoxin (ppb)	2387	2.200	0.4000	3.897	1.383	0.4614	45	-1.23	600.10	0
600	Total Aflatoxin (ppb)	2456	2.297	0.9880	3.897	1.383	0.4614	45	-1.16	600.99	0
600	Total Aflatoxin (ppb)	0006	2.300	0.3200	3.897	1.383	0.4614	45	-1.15	600.20	0
600	Total Aflatoxin (ppb)	0682	2.300	0.0000	3.897	1.383	0.4614	45	-1.15	600.24	0
600	Total Aflatoxin (ppb)	0353	2.350	0.0920	3.897	1.383	0.4614	45	-1.12	600.24	0
600	Total Aflatoxin (ppb)	0010	2.350	0.3000	3.897	1.383	0.4614	45	-1.12	600.20	0
600	Total Aflatoxin (ppb)	2369	2.470	0.2800	3.897	1.383	0.4614	45	-1.03	600.14	0
600	Total Aflatoxin (ppb)	2369	2.470	0.2800	3.897	1.383	0.4614	45	-1.03	600.20	0
600	Total Aflatoxin (ppb)	2443	2.560	0.0000	3.897	1.383	0.4614	45	-0.97	600.24	0
600	Total Aflatoxin (ppb)	2269	2.650	0.1000	3.897	1.383	0.4614	45	-0.90	600.98	0
600	Total Aflatoxin (ppb)	0511	2.700	0.4000	3.897	1.383	0.4614	45	-0.87	600.01	0
600	Total Aflatoxin (ppb)	0619	2.825	0.7300	3.897	1.383	0.4614	45	-0.77	600.24	0
600	Total Aflatoxin (ppb)	0001	2.890	0.1600	3.897	1.383	0.4614	45	-0.73	600.20	0
600	Total Aflatoxin (ppb)	0004	3.000	0.2000	3.897	1.383	0.4614	45	-0.65	600.14	0
600	Total Aflatoxin (ppb)	2096	3.350	0.3000	3.897	1.383	0.4614	45	-0.40	600.01	0
600	Total Aflatoxin (ppb)	0297	3.550	0.7000	3.897	1.383	0.4614	45	-0.25	600.24	0
600	Total Aflatoxin (ppb)	0027	3.590	0.7600	3.897	1.383	0.4614	45	-0.22	600.01	0
600	Total Aflatoxin (ppb)	0148	3.600	1.200	3.897	1.383	0.4614	45	-0.21	600.24	0
600	Total Aflatoxin (ppb)	0529	3.650	0.1000	3.897	1.383	0.4614	45	-0.18	600.01	0
600	Total Aflatoxin (ppb)	0629	3.860	0.1200	3.897	1.383	0.4614	45	-0.03	600.20	0
600	Total Aflatoxin (ppb)	0656	3.920	0.0000	3.897	1.383	0.4614	45	0.02	600.12	0
600	Total Aflatoxin (ppb)	0098	4.085	0.8900	3.897	1.383	0.4614	45	0.14	600.22	0
600	Total Aflatoxin (ppb)	2129	4.350	0.3000	3.897	1.383	0.4614	45	0.33	600.24	0
600	Total Aflatoxin (ppb)	0066	4.460	0.2800	3.897	1.383	0.4614	45	0.41	600.13	0
600	Total Aflatoxin (ppb)	2302	4.500	1.000	3.897	1.383	0.4614	45	0.44	600.99	0
600	Total Aflatoxin (ppb)	0034	4.925	0.7500	3.897	1.383	0.4614	45	0.74	600.01	0
600	Total Aflatoxin (ppb)	0590	4.945	0.7500	3.897	1.383	0.4614	45	0.76	600.12	0
600	Total Aflatoxin (ppb)	2226	5.142	0.0930	3.897	1.383	0.4614	45	0.90	600.12	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT	Your	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests	Z Score	Method	
600	Total Aflatoxin (ppb)	0554	5.150	0.3000	3.897	1.383	0.4614	45	0.91	600.01	0
600	Total Aflatoxin (ppb)	0354	5.220	0.1200	3.897	1.383	0.4614	45	0.96	600.08	0
600	Total Aflatoxin (ppb)	0674	5.400	0.0000	3.897	1.383	0.4614	45	1.09	600.20	0
600	Total Aflatoxin (ppb)	0407	6.090	1.140	3.897	1.383	0.4614	45	1.59	600.13	0
600	Total Aflatoxin (ppb)	0596	6.200	0.4000	3.897	1.383	0.4614	45	1.67	600.98	0
600	Total Aflatoxin (ppb)	0354	6.280	0.3800	3.897	1.383	0.4614	45	1.72	600.14	0
600	Total Aflatoxin (ppb)	0354	6.280	0.3800	3.897	1.383	0.4614	45	1.72	600.21	0
600	Total Aflatoxin (ppb)	0354	6.280	0.3800	3.897	1.383	0.4614	45	1.72	600.22	0
600	Total Aflatoxin (ppb)	0227	6.500	1.000	3.897	1.383	0.4614	45	1.88	600.01	0
600	Total Aflatoxin (ppb)	2022	8.110	0.1800	3.897	1.383	0.4614	45	3.05	600.08	0
600	Total Aflatoxin (ppb)	2199	22.75	2.500	3.897	1.383	0.4614	45	13.63	600.14	0
600	Total Aflatoxin (ppb)	0042	24.37	1.360	3.897	1.383	0.4614	45	14.80	600.20	0
600	Total Aflatoxin (ppb)	0014	8.050	5.700	3.897	1.383	0.4614	45	3.00	600.01	1
600	Total Aflatoxin (ppb)	2326	< 0.5		3.897	1.383	0.4614	45		600.23	5
600	Total Aflatoxin (ppb)	2113	< 5		3.897	1.383	0.4614	45		600.01	5
600	Total Aflatoxin (ppb)	0563	< 5		3.897	1.383	0.4614	45		600.20	5
600	Total Aflatoxin (ppb)	2033	< 10		3.897	1.383	0.4614	45		600.24	5
601	Aflatoxin B1 (ppb)	2325	1.380	0.0200	3.027	1.082	0.3527	29	-1.52	601.24	0
601	Aflatoxin B1 (ppb)	0689	1.700	0.2000	3.027	1.082	0.3527	29	-1.23	601.16	0
601	Aflatoxin B1 (ppb)	2113	1.700	0.2000	3.027	1.082	0.3527	29	-1.23	601.23	0
601	Aflatoxin B1 (ppb)	0723	1.850	0.0000	3.027	1.082	0.3527	29	-1.09	601.20	0
601	Aflatoxin B1 (ppb)	2327	1.960	0.6200	3.027	1.082	0.3527	29	-0.99	601.23	0
601	Aflatoxin B1 (ppb)	0964	2.000	0.0000	3.027	1.082	0.3527	29	-0.95	601.23	0
601	Aflatoxin B1 (ppb)	2322	2.060	0.3600	3.027	1.082	0.3527	29	-0.89	601.20	0
601	Aflatoxin B1 (ppb)	0010	2.150	0.3000	3.027	1.082	0.3527	29	-0.81	601.20	0
601	Aflatoxin B1 (ppb)	0006	2.180	0.2800	3.027	1.082	0.3527	29	-0.78	601.20	0
601	Aflatoxin B1 (ppb)	0353	2.259	0.0580	3.027	1.082	0.3527	29	-0.71	601.23	0
601	Aflatoxin B1 (ppb)	2369	2.265	0.2300	3.027	1.082	0.3527	29	-0.70	601.20	0
601	Aflatoxin B1 (ppb)	2369	2.265	0.2300	3.027	1.082	0.3527	29	-0.70	601.24	0
601	Aflatoxin B1 (ppb)	0682	2.300	0.0000	3.027	1.082	0.3527	29	-0.67	601.23	0
601	Aflatoxin B1 (ppb)	2443	2.560	0.0000	3.027	1.082	0.3527	29	-0.43	601.23	0
601	Aflatoxin B1 (ppb)	0009	2.807	0.2250	3.027	1.082	0.3527	29	-0.20	601.23	0
601	Aflatoxin B1 (ppb)	0619	2.825	0.7300	3.027	1.082	0.3527	29	-0.19	601.23	0
601	Aflatoxin B1 (ppb)	2192	2.990	0.1800	3.027	1.082	0.3527	29	-0.03	601.20	0
601	Aflatoxin B1 (ppb)	0675	3.535	0.2900	3.027	1.082	0.3527	29	0.47	601.20	0
601	Aflatoxin B1 (ppb)	0297	3.550	0.7000	3.027	1.082	0.3527	29	0.48	601.23	0
601	Aflatoxin B1 (ppb)	0148	3.600	1.200	3.027	1.082	0.3527	29	0.53	601.23	0
601	Aflatoxin B1 (ppb)	0629	3.605	0.1300	3.027	1.082	0.3527	29	0.53	601.20	0
601	Aflatoxin B1 (ppb)	2302	4.000	2.000	3.027	1.082	0.3527	29	0.90	601.99	0
601	Aflatoxin B1 (ppb)	0098	4.085	0.8900	3.027	1.082	0.3527	29	0.98	601.24	0
601	Aflatoxin B1 (ppb)	2269	4.100	0.2000	3.027	1.082	0.3527	29	0.99	601.98	0
601	Aflatoxin B1 (ppb)	0028	4.435	0.3300	3.027	1.082	0.3527	29	1.30	601.23	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT	Your	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests	Z Score	Method	
601	Aflatoxin B1 (ppb)	0505	4.805	0.3300	3.027	1.082	0.3527	29	1.64	601.16	0
601	Aflatoxin B1 (ppb)	0354	5.925	0.1900	3.027	1.082	0.3527	29	2.68	601.20	0
601	Aflatoxin B1 (ppb)	0354	5.925	0.1900	3.027	1.082	0.3527	29	2.68	601.24	0
601	Aflatoxin B1 (ppb)	2022	7.435	0.5700	3.027	1.082	0.3527	29	4.08	601.91	0
601	Aflatoxin B1 (ppb)	2326	< 0.5		3.027	1.082	0.3527	29		601.23	5
601	Aflatoxin B1 (ppb)	0242	< 1		3.027	1.082	0.3527	29		601.23	5
601	Aflatoxin B1 (ppb)	0870	< 2		3.027	1.082	0.3527	29		601.23	5
601	Aflatoxin B1 (ppb)	0563	< 5		3.027	1.082	0.3527	29		601.20	5
601	Aflatoxin B1 (ppb)	0572	< 10		3.027	1.082	0.3527	29		601.23	5
601	Aflatoxin B1 (ppb)	2033	< 10		3.027	1.082	0.3527	29		601.23	5
601	Aflatoxin B1 (ppb)	2140	< 20		3.027	1.082	0.3527	29		601.23	5
602	Aflatoxin B2 (ppb)	0006	0.1150	0.0500	0.2461	0.0941	0.0883	9	-1.39	602.20	0
602	Aflatoxin B2 (ppb)	0353	0.1700	0.0080	0.2461	0.0941	0.0883	9	-0.81	602.23	0
602	Aflatoxin B2 (ppb)	0723	0.1700	0.0000	0.2461	0.0941	0.0883	9	-0.81	602.20	0
602	Aflatoxin B2 (ppb)	0010	0.2000	0.0000	0.2461	0.0941	0.0883	9	-0.49	602.20	0
602	Aflatoxin B2 (ppb)	2369	0.2000	0.0400	0.2461	0.0941	0.0883	9	-0.49	602.20	0
602	Aflatoxin B2 (ppb)	0629	0.2550	0.0100	0.2461	0.0941	0.0883	9	0.09	602.20	0
602	Aflatoxin B2 (ppb)	0354	0.3550	0.1900	0.2461	0.0941	0.0883	9	1.16	602.20	0
602	Aflatoxin B2 (ppb)	0354	0.3550	0.1900	0.2461	0.0941	0.0883	9	1.16	602.24	0
602	Aflatoxin B2 (ppb)	0675	0.3950	0.1300	0.2461	0.0941	0.0883	9	1.58	602.20	0
602	Aflatoxin B2 (ppb)	0619	0.0000	0.0000	0.2461	0.0941	0.0883	9		602.23	4
602	Aflatoxin B2 (ppb)	2325	0.0000	0.0000	0.2461	0.0941	0.0883	9		602.24	4
602	Aflatoxin B2 (ppb)	0297	< 0.3		0.2461	0.0941	0.0883	9		602.23	5
602	Aflatoxin B2 (ppb)	0682	< 0.5		0.2461	0.0941	0.0883	9		602.23	5
602	Aflatoxin B2 (ppb)	2326	< 0.5		0.2461	0.0941	0.0883	9		602.23	5
602	Aflatoxin B2 (ppb)	0870	< 0.6		0.2461	0.0941	0.0883	9		602.23	5
602	Aflatoxin B2 (ppb)	0009	< 0.63		0.2461	0.0941	0.0883	9		602.23	5
602	Aflatoxin B2 (ppb)	0242	< 1		0.2461	0.0941	0.0883	9		602.23	5
602	Aflatoxin B2 (ppb)	0964	< 1		0.2461	0.0941	0.0883	9		602.23	5
602	Aflatoxin B2 (ppb)	2113	< 1		0.2461	0.0941	0.0883	9		602.23	5
602	Aflatoxin B2 (ppb)	0098	< 1		0.2461	0.0941	0.0883	9		602.24	5
602	Aflatoxin B2 (ppb)	2327	< 1.2		0.2461	0.0941	0.0883	9		602.23	5
602	Aflatoxin B2 (ppb)	0148	< 2.5		0.2461	0.0941	0.0883	9		602.23	5
602	Aflatoxin B2 (ppb)	0563	< 5		0.2461	0.0941	0.0883	9		602.20	5
602	Aflatoxin B2 (ppb)	2033	< 10		0.2461	0.0941	0.0883	9		602.23	5
602	Aflatoxin B2 (ppb)	2140	< 20		0.2461	0.0941	0.0883	9		602.23	5
603	Aflatoxin G1 (ppb)	0723	0.1200	0.0000				3		603.20	0
603	Aflatoxin G1 (ppb)	0964	1.000	0.0000				3		603.23	0
603	Aflatoxin G1 (ppb)	0505	1.140	0.0600				3		603.16	0
603	Aflatoxin G1 (ppb)	2302	1.500	1.000				3		603.99	1
603	Aflatoxin G1 (ppb)	0675	0.0000	0.0000				3		603.20	4
603	Aflatoxin G1 (ppb)	0297	0.0000	0.0000				3		603.23	4

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests			
603	Aflatoxin G1 (ppb)	0619	0.0000	0.0000				3	603.23	4	
603	Aflatoxin G1 (ppb)	2325	0.0000	0.0000				3	603.24	4	
603	Aflatoxin G1 (ppb)	0006	< 0.05					3	603.20	5	
603	Aflatoxin G1 (ppb)	0629	< 0.25					3	603.20	5	
603	Aflatoxin G1 (ppb)	0682	< 0.5					3	603.23	5	
603	Aflatoxin G1 (ppb)	2326	< 0.5					3	603.23	5	
603	Aflatoxin G1 (ppb)	0870	< 0.6					3	603.23	5	
603	Aflatoxin G1 (ppb)	0009	< 0.69					3	603.23	5	
603	Aflatoxin G1 (ppb)	0242	< 1					3	603.23	5	
603	Aflatoxin G1 (ppb)	2327	< 1.1					3	603.23	5	
603	Aflatoxin G1 (ppb)	0148	< 2.5					3	603.23	5	
603	Aflatoxin G1 (ppb)	2033	< 10					3	603.23	5	
603	Aflatoxin G1 (ppb)	0098	< 15					3	603.24	5	
603	Aflatoxin G1 (ppb)	2140	< 20					3	603.23	5	
604	Aflatoxin G2 (ppb)	0723	0.0200	0.0000				2	604.20	0	
604	Aflatoxin G2 (ppb)	0964	1.000	0.0000				2	604.23	0	
604	Aflatoxin G2 (ppb)	0675	0.0000	0.0000				2	604.20	4	
604	Aflatoxin G2 (ppb)	0297	0.0000	0.0000				2	604.23	4	
604	Aflatoxin G2 (ppb)	0619	0.0000	0.0000				2	604.23	4	
604	Aflatoxin G2 (ppb)	2325	0.0000	0.0000				2	604.24	4	
604	Aflatoxin G2 (ppb)	0006	< 0.05					2	604.20	5	
604	Aflatoxin G2 (ppb)	0629	< 0.25					2	604.20	5	
604	Aflatoxin G2 (ppb)	0682	< 0.5					2	604.23	5	
604	Aflatoxin G2 (ppb)	2326	< 0.5					2	604.23	5	
604	Aflatoxin G2 (ppb)	0870	< 0.6					2	604.23	5	
604	Aflatoxin G2 (ppb)	0009	< 0.69					2	604.23	5	
604	Aflatoxin G2 (ppb)	0242	< 1					2	604.23	5	
604	Aflatoxin G2 (ppb)	2327	< 1.6					2	604.23	5	
604	Aflatoxin G2 (ppb)	0148	< 3					2	604.23	5	
604	Aflatoxin G2 (ppb)	0098	< 5					2	604.24	5	
604	Aflatoxin G2 (ppb)	2033	< 10					2	604.23	5	
604	Aflatoxin G2 (ppb)	2140	< 20					2	604.23	5	
610	Deoxynivalenol (ppb)	0407	2.057	0.0890	2,203	658.4	207.4	53	-3.34	610.07	0
610	Deoxynivalenol (ppb)	0554	2.650	0.7000	2,203	658.4	207.4	53	-3.34	610.01	0
610	Deoxynivalenol (ppb)	0596	3.000	0.0000	2,203	658.4	207.4	53	-3.34	610.98	0
610	Deoxynivalenol (ppb)	2268	280.5	7.000	2,203	658.4	207.4	53	-2.92	610.22	0
610	Deoxynivalenol (ppb)	0028	1,417	353.6	2,203	658.4	207.4	53	-1.19	610.23	0
610	Deoxynivalenol (ppb)	2345	1,515	570.0	2,203	658.4	207.4	53	-1.04	610.09	0
610	Deoxynivalenol (ppb)	0619	1,522	244.0	2,203	658.4	207.4	53	-1.03	610.23	0
610	Deoxynivalenol (ppb)	0689	1,537	52.00	2,203	658.4	207.4	53	-1.01	610.16	0
610	Deoxynivalenol (ppb)	0964	1,588	323.0	2,203	658.4	207.4	53	-0.93	610.23	0
610	Deoxynivalenol (ppb)	0572	1,600	200.0	2,203	658.4	207.4	53	-0.92	610.23	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT	Your	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests	Z Score	Method	
610	Deoxynivalenol (ppb)	2129	1,690	110.0	2,203	658.4	207.4	53	-0.78	610.23	0
610	Deoxynivalenol (ppb)	2456	1,744	204.8	2,203	658.4	207.4	53	-0.70	610.99	0
610	Deoxynivalenol (ppb)	2055	1,745	31.00	2,203	658.4	207.4	53	-0.70	610.23	0
610	Deoxynivalenol (ppb)	2326	1,838	19.00	2,203	658.4	207.4	53	-0.55	610.23	0
610	Deoxynivalenol (ppb)	0047	1,858	448.3	2,203	658.4	207.4	53	-0.52	610.16	0
610	Deoxynivalenol (ppb)	0353	1,881	1.000	2,203	658.4	207.4	53	-0.49	610.23	0
610	Deoxynivalenol (ppb)	2033	1,984	24.00	2,203	658.4	207.4	53	-0.33	610.23	0
610	Deoxynivalenol (ppb)	0148	2,018	200.7	2,203	658.4	207.4	53	-0.28	610.23	0
610	Deoxynivalenol (ppb)	0682	2,050	0.0000	2,203	658.4	207.4	53	-0.23	610.23	0
610	Deoxynivalenol (ppb)	2369	2,060	15.00	2,203	658.4	207.4	53	-0.22	610.09	0
610	Deoxynivalenol (ppb)	2369	2,060	15.00	2,203	658.4	207.4	53	-0.22	610.20	0
610	Deoxynivalenol (ppb)	2140	2,124	164.0	2,203	658.4	207.4	53	-0.12	610.23	0
610	Deoxynivalenol (ppb)	0010	2,141	68.00	2,203	658.4	207.4	53	-0.09	610.23	0
610	Deoxynivalenol (ppb)	0208	2,190	860.0	2,203	658.4	207.4	53	-0.02	610.01	0
610	Deoxynivalenol (ppb)	2302	2,193	321.0	2,203	658.4	207.4	53	-0.02	610.99	0
610	Deoxynivalenol (ppb)	0242	2,206	652.0	2,203	658.4	207.4	53	0.00	610.23	0
610	Deoxynivalenol (ppb)	0674	2,227	0.0000	2,203	658.4	207.4	53	0.04	610.20	0
610	Deoxynivalenol (ppb)	0042	2,250	300.0	2,203	658.4	207.4	53	0.07	610.01	0
610	Deoxynivalenol (ppb)	2269	2,280	240.1	2,203	658.4	207.4	53	0.12	610.98	0
610	Deoxynivalenol (ppb)	0098	2,300	0.0000	2,203	658.4	207.4	53	0.15	610.20	0
610	Deoxynivalenol (ppb)	2327	2,357	38.00	2,203	658.4	207.4	53	0.23	610.23	0
610	Deoxynivalenol (ppb)	0590	2,360	180.0	2,203	658.4	207.4	53	0.24	610.07	0
610	Deoxynivalenol (ppb)	0006	2,424	222.0	2,203	658.4	207.4	53	0.34	610.20	0
610	Deoxynivalenol (ppb)	0034	2,425	150.0	2,203	658.4	207.4	53	0.34	610.01	0
610	Deoxynivalenol (ppb)	0027	2,430	160.0	2,203	658.4	207.4	53	0.35	610.01	0
610	Deoxynivalenol (ppb)	0870	2,434	24.90	2,203	658.4	207.4	53	0.35	610.99	0
610	Deoxynivalenol (ppb)	0870	2,440	40.00	2,203	658.4	207.4	53	0.36	610.23	0
610	Deoxynivalenol (ppb)	2192	2,450	0.0000	2,203	658.4	207.4	53	0.38	610.01	0
610	Deoxynivalenol (ppb)	0009	2,470	416.3	2,203	658.4	207.4	53	0.41	610.23	0
610	Deoxynivalenol (ppb)	2113	2,500	600.0	2,203	658.4	207.4	53	0.45	610.23	0
610	Deoxynivalenol (ppb)	0629	2,512	144.0	2,203	658.4	207.4	53	0.47	610.07	0
610	Deoxynivalenol (ppb)	0297	2,590	340.0	2,203	658.4	207.4	53	0.59	610.23	0
610	Deoxynivalenol (ppb)	0227	2,600	200.0	2,203	658.4	207.4	53	0.60	610.01	0
610	Deoxynivalenol (ppb)	0066	2,695	390.0	2,203	658.4	207.4	53	0.75	610.08	0
610	Deoxynivalenol (ppb)	0511	2,700	200.0	2,203	658.4	207.4	53	0.76	610.01	0
610	Deoxynivalenol (ppb)	0354	2,759	48.48	2,203	658.4	207.4	53	0.85	610.05	0
610	Deoxynivalenol (ppb)	0505	2,977	238.6	2,203	658.4	207.4	53	1.18	610.16	0
610	Deoxynivalenol (ppb)	0656	2,989	0.0000	2,203	658.4	207.4	53	1.19	610.07	0
610	Deoxynivalenol (ppb)	2113	3,000	400.0	2,203	658.4	207.4	53	1.21	610.01	0
610	Deoxynivalenol (ppb)	0014	3,100	800.0	2,203	658.4	207.4	53	1.36	610.01	0
610	Deoxynivalenol (ppb)	0035	3,139	70.00	2,203	658.4	207.4	53	1.42	610.08	0
610	Deoxynivalenol (ppb)	0529	3,250	700.0	2,203	658.4	207.4	53	1.59	610.01	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT	Your	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests	Z Score	Method	
610	Deoxynivalenol (ppb)	2022	3,300	200.0	2,203	658.4	207.4	53	1.67	610.05	0
620	Total Fumonisin (ppb)	2268	1.650	0.3000	1,776	534.0	128.7	30	-3.32	620.11	0
620	Total Fumonisin (ppb)	0098	1.700	0.2000	1,776	534.0	128.7	30	-3.32	620.11	0
620	Total Fumonisin (ppb)	0596	2.200	0.2000	1,776	534.0	128.7	30	-3.32	620.98	0
620	Total Fumonisin (ppb)	0554	2.700	1.200	1,776	534.0	128.7	30	-3.32	620.01	0
620	Total Fumonisin (ppb)	0505	384.3	6.170	1,776	534.0	128.7	30	-2.61	620.16	0
620	Total Fumonisin (ppb)	2302	542.5	327.0	1,776	534.0	128.7	30	-2.31	620.99	0
620	Total Fumonisin (ppb)	0656	1,232	0.0000	1,776	534.0	128.7	30	-1.02	620.09	0
620	Total Fumonisin (ppb)	2443	1,379	0.0000	1,776	534.0	128.7	30	-0.74	620.23	0
620	Total Fumonisin (ppb)	2369	1,439	14.00	1,776	534.0	128.7	30	-0.63	620.11	0
620	Total Fumonisin (ppb)	0619	1,622	227.0	1,776	534.0	128.7	30	-0.29	620.23	0
620	Total Fumonisin (ppb)	0674	1,740	0.0000	1,776	534.0	128.7	30	-0.07	620.07	0
620	Total Fumonisin (ppb)	2456	1,916	39.19	1,776	534.0	128.7	30	0.26	620.99	0
620	Total Fumonisin (ppb)	0682	1,920	0.0000	1,776	534.0	128.7	30	0.27	620.23	0
620	Total Fumonisin (ppb)	0010	1,935	10.00	1,776	534.0	128.7	30	0.30	620.23	0
620	Total Fumonisin (ppb)	0407	1,953	75.00	1,776	534.0	128.7	30	0.33	620.09	0
620	Total Fumonisin (ppb)	0227	2,000	0.0000	1,776	534.0	128.7	30	0.42	620.01	0
620	Total Fumonisin (ppb)	0511	2,000	200.0	1,776	534.0	128.7	30	0.42	620.01	0
620	Total Fumonisin (ppb)	0035	2,030	80.00	1,776	534.0	128.7	30	0.48	620.09	0
620	Total Fumonisin (ppb)	0042	2,050	300.0	1,776	534.0	128.7	30	0.51	620.01	0
620	Total Fumonisin (ppb)	2129	2,066	204.0	1,776	534.0	128.7	30	0.54	620.23	0
620	Total Fumonisin (ppb)	0014	2,130	420.0	1,776	534.0	128.7	30	0.66	620.01	0
620	Total Fumonisin (ppb)	0529	2,350	500.0	1,776	534.0	128.7	30	1.07	620.02	0
620	Total Fumonisin (ppb)	2226	2,388	39.00	1,776	534.0	128.7	30	1.14	620.09	0
620	Total Fumonisin (ppb)	0027	2,396	342.0	1,776	534.0	128.7	30	1.16	620.22	0
620	Total Fumonisin (ppb)	2113	2,400	200.0	1,776	534.0	128.7	30	1.17	620.01	0
620	Total Fumonisin (ppb)	0353	2,461	30.00	1,776	534.0	128.7	30	1.28	620.23	0
620	Total Fumonisin (ppb)	0066	2,575	50.00	1,776	534.0	128.7	30	1.50	620.09	0
620	Total Fumonisin (ppb)	2033	2,634	96.00	1,776	534.0	128.7	30	1.61	620.23	0
620	Total Fumonisin (ppb)	0208	2,650	100.0	1,776	534.0	128.7	30	1.64	620.98	0
620	Total Fumonisin (ppb)	2022	5,775	190.0	1,776	534.0	128.7	30	7.49	620.07	0
620	Total Fumonisin (ppb)	0689	< 200		1,776	534.0	128.7	30		620.16	5
621	Fumonsin B1 (ppb)	2443	1,075	0.0000	1,600	482.5	133.0	17	-1.09	621.23	0
621	Fumonsin B1 (ppb)	0964	1,078	553.5	1,600	482.5	133.0	17	-1.08	621.23	0
621	Fumonsin B1 (ppb)	2369	1,242	12.00	1,600	482.5	133.0	17	-0.74	621.21	0
621	Fumonsin B1 (ppb)	0009	1,247	179.1	1,600	482.5	133.0	17	-0.73	621.23	0
621	Fumonsin B1 (ppb)	0619	1,258	198.0	1,600	482.5	133.0	17	-0.71	621.23	0
621	Fumonsin B1 (ppb)	0098	1,400	200.0	1,600	482.5	133.0	17	-0.42	621.21	0
621	Fumonsin B1 (ppb)	0010	1,495	50.00	1,600	482.5	133.0	17	-0.22	621.23	0
621	Fumonsin B1 (ppb)	2326	1,646	267.0	1,600	482.5	133.0	17	0.09	621.23	0
621	Fumonsin B1 (ppb)	0148	1,652	65.55	1,600	482.5	133.0	17	0.11	621.23	0
621	Fumonsin B1 (ppb)	0682	1,670	0.0000	1,600	482.5	133.0	17	0.14	621.23	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT	Your	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests	Z Score	Method	
621	Fumonsin B1 (ppb)	2140	1,753	390.0	1,600	482.5	133.0	17	0.32	621.23	0
621	Fumonsin B1 (ppb)	0353	1,756	3.000	1,600	482.5	133.0	17	0.32	621.23	0
621	Fumonsin B1 (ppb)	0870	1,780	40.00	1,600	482.5	133.0	17	0.37	621.23	0
621	Fumonsin B1 (ppb)	2033	1,925	70.00	1,600	482.5	133.0	17	0.67	621.23	0
621	Fumonsin B1 (ppb)	0027	2,022	89.00	1,600	482.5	133.0	17	0.87	621.22	0
621	Fumonsin B1 (ppb)	2055	2,052	72.00	1,600	482.5	133.0	17	0.94	621.23	0
621	Fumonsin B1 (ppb)	2327	2,155	127.0	1,600	482.5	133.0	17	1.15	621.23	0
621	Fumonsin B1 (ppb)	0572	< 500		1,600	482.5	133.0	17		621.23	5
622	Fumonisin B2 (ppb)	2369	197.5	3.000	411.7	128.8	36.92	17	-1.66	622.21	0
622	Fumonisin B2 (ppb)	0682	250.0	0.0000	411.7	128.8	36.92	17	-1.26	622.23	0
622	Fumonisin B2 (ppb)	0098	300.0	0.0000	411.7	128.8	36.92	17	-0.87	622.21	0
622	Fumonisin B2 (ppb)	2443	303.9	0.0000	411.7	128.8	36.92	17	-0.84	622.23	0
622	Fumonisin B2 (ppb)	0010	310.0	20.00	411.7	128.8	36.92	17	-0.79	622.23	0
622	Fumonisin B2 (ppb)	2326	349.0	64.00	411.7	128.8	36.92	17	-0.49	622.23	0
622	Fumonisin B2 (ppb)	0619	363.5	29.00	411.7	128.8	36.92	17	-0.37	622.23	0
622	Fumonisin B2 (ppb)	0027	374.5	253.0	411.7	128.8	36.92	17	-0.29	622.22	0
622	Fumonisin B2 (ppb)	2140	388.0	30.00	411.7	128.8	36.92	17	-0.18	622.23	0
622	Fumonisin B2 (ppb)	0870	430.0	20.00	411.7	128.8	36.92	17	0.14	622.23	0
622	Fumonisin B2 (ppb)	0964	489.3	85.50	411.7	128.8	36.92	17	0.60	622.23	0
622	Fumonisin B2 (ppb)	2327	515.5	17.00	411.7	128.8	36.92	17	0.81	622.23	0
622	Fumonisin B2 (ppb)	0009	524.8	28.90	411.7	128.8	36.92	17	0.88	622.23	0
622	Fumonisin B2 (ppb)	0353	540.5	13.00	411.7	128.8	36.92	17	1.00	622.23	0
622	Fumonisin B2 (ppb)	2055	542.0	234.0	411.7	128.8	36.92	17	1.01	622.23	0
622	Fumonisin B2 (ppb)	2033	546.5	19.00	411.7	128.8	36.92	17	1.05	622.23	0
622	Fumonisin B2 (ppb)	0148	554.4	5.830	411.7	128.8	36.92	17	1.11	622.23	0
623	Fumonisin B3 (ppb)	0010	130.0	40.00	174.8	55.95	31.52	8	-0.80	623.23	0
623	Fumonisin B3 (ppb)	0353	130.5	49.00	174.8	55.95	31.52	8	-0.79	623.23	0
623	Fumonisin B3 (ppb)	2033	162.5	7.000	174.8	55.95	31.52	8	-0.22	623.23	0
623	Fumonisin B3 (ppb)	2326	166.0	28.00	174.8	55.95	31.52	8	-0.16	623.23	0
623	Fumonisin B3 (ppb)	2327	181.0	14.00	174.8	55.95	31.52	8	0.11	623.23	0
623	Fumonisin B3 (ppb)	0009	192.9	19.16	174.8	55.95	31.52	8	0.32	623.23	0
623	Fumonisin B3 (ppb)	0870	200.0	0.0000	174.8	55.95	31.52	8	0.45	623.23	0
623	Fumonisin B3 (ppb)	0964	274.3	63.50	174.8	55.95	31.52	8	1.78	623.23	0
630	Ochratoxin A (ppb)	2327	1.525	0.1300	4.048	1.435	0.3473	27	-1.76	630.23	0
630	Ochratoxin A (ppb)	0629	2.200	0.1000	4.048	1.435	0.3473	27	-1.29	630.20	0
630	Ochratoxin A (ppb)	0870	2.280	0.3600	4.048	1.435	0.3473	27	-1.23	630.23	0
630	Ochratoxin A (ppb)	0563	2.500	0.6000	4.048	1.435	0.3473	27	-1.08	630.20	0
630	Ochratoxin A (ppb)	2325	2.665	0.0100	4.048	1.435	0.3473	27	-0.96	630.20	0
630	Ochratoxin A (ppb)	2456	2.679	0.3340	4.048	1.435	0.3473	27	-0.95	630.99	0
630	Ochratoxin A (ppb)	2226	2.727	0.4820	4.048	1.435	0.3473	27	-0.92	630.05	0
630	Ochratoxin A (ppb)	0505	2.875	0.0100	4.048	1.435	0.3473	27	-0.82	630.16	0
630	Ochratoxin A (ppb)	0353	2.884	0.4620	4.048	1.435	0.3473	27	-0.81	630.23	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT	Your	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests	Z Score	Method	
630	Ochratoxin A (ppb)	0619	2.895	0.6900	4.048	1.435	0.3473	27	-0.80	630.23	0
630	Ochratoxin A (ppb)	0675	2.940	0.0200	4.048	1.435	0.3473	27	-0.77	630.20	0
630	Ochratoxin A (ppb)	0407	2.950	0.6000	4.048	1.435	0.3473	27	-0.76	630.05	0
630	Ochratoxin A (ppb)	2322	3.155	0.4300	4.048	1.435	0.3473	27	-0.62	630.20	0
630	Ochratoxin A (ppb)	2369	3.290	0.1200	4.048	1.435	0.3473	27	-0.53	630.06	0
630	Ochratoxin A (ppb)	2369	3.290	0.1200	4.048	1.435	0.3473	27	-0.53	630.20	0
630	Ochratoxin A (ppb)	0682	3.300	0.0000	4.048	1.435	0.3473	27	-0.52	630.23	0
630	Ochratoxin A (ppb)	0354	3.400	0.1600	4.048	1.435	0.3473	27	-0.45	630.03	0
630	Ochratoxin A (ppb)	0723	4.040	0.0000	4.048	1.435	0.3473	27	-0.01	630.20	0
630	Ochratoxin A (ppb)	0066	4.200	0.2000	4.048	1.435	0.3473	27	0.11	630.05	0
630	Ochratoxin A (ppb)	0148	5.375	1.190	4.048	1.435	0.3473	27	0.92	630.23	0
630	Ochratoxin A (ppb)	0596	5.800	0.4000	4.048	1.435	0.3473	27	1.22	630.98	0
630	Ochratoxin A (ppb)	2129	6.250	0.1000	4.048	1.435	0.3473	27	1.53	630.23	0
630	Ochratoxin A (ppb)	2443	7.620	0.0000	4.048	1.435	0.3473	27	2.49	630.23	0
630	Ochratoxin A (ppb)	2022	9.595	0.7100	4.048	1.435	0.3473	27	3.87	630.03	0
630	Ochratoxin A (ppb)	0674	13.64	0.0000	4.048	1.435	0.3473	27	6.68	630.20	0
630	Ochratoxin A (ppb)	0529	13.70	2.000	4.048	1.435	0.3473	27	6.73	630.01	0
630	Ochratoxin A (ppb)	0656	25.73	0.0000	4.048	1.435	0.3473	27	15.11	630.05	0
630	Ochratoxin A (ppb)	0354	0.0000	0.0000	4.048	1.435	0.3473	27		630.06	4
630	Ochratoxin A (ppb)	0354	0.0000	0.0000	4.048	1.435	0.3473	27		630.20	4
630	Ochratoxin A (ppb)	2302	1.000	2.000	4.048	1.435	0.3473	27		630.99	4
630	Ochratoxin A (ppb)	2326	< 0.5		4.048	1.435	0.3473	27		630.23	5
630	Ochratoxin A (ppb)	0964	< 1		4.048	1.435	0.3473	27		630.23	5
630	Ochratoxin A (ppb)	0227	< 2		4.048	1.435	0.3473	27		630.01	5
630	Ochratoxin A (ppb)	0689	< 2.5		4.048	1.435	0.3473	27		630.16	5
630	Ochratoxin A (ppb)	2033	< 5		4.048	1.435	0.3473	27		630.23	5
630	Ochratoxin A (ppb)	0098	< 10		4.048	1.435	0.3473	27		630.20	5
630	Ochratoxin A (ppb)	0572	< 20		4.048	1.435	0.3473	27		630.23	5
630	Ochratoxin A (ppb)	2140	< 20		4.048	1.435	0.3473	27		630.23	5
640	T-2 Toxin (ppb)	2326	55.50	3.000	111.4	36.10	9.404	20	-1.55	640.23	0
640	T-2 Toxin (ppb)	2129	56.55	1.100	111.4	36.10	9.404	20	-1.52	640.23	0
640	T-2 Toxin (ppb)	0297	65.00	16.00	111.4	36.10	9.404	20	-1.29	640.23	0
640	T-2 Toxin (ppb)	0638	74.00	10.00	111.4	36.10	9.404	20	-1.04	640.05	0
640	T-2 Toxin (ppb)	0682	87.70	0.0000	111.4	36.10	9.404	20	-0.66	640.23	0
640	T-2 Toxin (ppb)	2055	89.25	4.500	111.4	36.10	9.404	20	-0.61	640.23	0
640	T-2 Toxin (ppb)	0964	92.50	109.0	111.4	36.10	9.404	20	-0.52	640.23	0
640	T-2 Toxin (ppb)	2033	100.0	0.0000	111.4	36.10	9.404	20	-0.32	640.23	0
640	T-2 Toxin (ppb)	0674	106.4	0.0000	111.4	36.10	9.404	20	-0.14	640.03	0
640	T-2 Toxin (ppb)	2140	111.0	4.000	111.4	36.10	9.404	20	-0.01	640.23	0
640	T-2 Toxin (ppb)	0870	116.8	32.46	111.4	36.10	9.404	20	0.15	640.23	0
640	T-2 Toxin (ppb)	2327	119.5	7.000	111.4	36.10	9.404	20	0.22	640.23	0
640	T-2 Toxin (ppb)	0619	121.5	1.000	111.4	36.10	9.404	20	0.28	640.23	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT	Your	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests	Z Score	Method	
640	T-2 Toxin (ppb)	0148	122.9	4.430	111.4	36.10	9.404	20	0.32	640.23	0
640	T-2 Toxin (ppb)	2369	131.1	4.170	111.4	36.10	9.404	20	0.55	640.23	0
640	T-2 Toxin (ppb)	0353	139.0	4.000	111.4	36.10	9.404	20	0.76	640.23	0
640	T-2 Toxin (ppb)	0407	142.3	6.095	111.4	36.10	9.404	20	0.85	640.05	0
640	T-2 Toxin (ppb)	0066	156.7	15.63	111.4	36.10	9.404	20	1.25	640.03	0
640	T-2 Toxin (ppb)	2269	290.8	37.30	111.4	36.10	9.404	20	4.97	640.98	0
640	T-2 Toxin (ppb)	0656	429.7	0.0000	111.4	36.10	9.404	20	8.82	640.05	0
640	T-2 Toxin (ppb)	0572	< 100		111.4	36.10	9.404	20		640.23	5
641	HT-2 Toxin (ppb)	2326	83.50	1.000	144.0	46.34	10.91	11	-1.31	641.23	0
641	HT-2 Toxin (ppb)	0297	109.9	12.80	144.0	46.34	10.91	11	-0.74	641.23	0
641	HT-2 Toxin (ppb)	2033	133.0	4.000	144.0	46.34	10.91	11	-0.24	641.23	0
641	HT-2 Toxin (ppb)	0619	134.5	1.000	144.0	46.34	10.91	11	-0.20	641.23	0
641	HT-2 Toxin (ppb)	2369	135.6	7.460	144.0	46.34	10.91	11	-0.18	641.23	0
641	HT-2 Toxin (ppb)	2327	138.0	6.000	144.0	46.34	10.91	11	-0.13	641.23	0
641	HT-2 Toxin (ppb)	0964	140.0	50.00	144.0	46.34	10.91	11	-0.09	641.23	0
641	HT-2 Toxin (ppb)	2140	161.5	47.00	144.0	46.34	10.91	11	0.38	641.23	0
641	HT-2 Toxin (ppb)	0148	173.9	0.4800	144.0	46.34	10.91	11	0.65	641.23	0
641	HT-2 Toxin (ppb)	0870	175.1	9.200	144.0	46.34	10.91	11	0.67	641.23	0
641	HT-2 Toxin (ppb)	0353	185.0	20.00	144.0	46.34	10.91	11	0.89	641.23	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0689	108.5	5.000	283.2	89.50	22.04	15	-1.95	642.16	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0297	175.1	2.590	283.2	89.50	22.04	15	-1.21	642.04	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2302	181.5	7.000	283.2	89.50	22.04	15	-1.14	642.99	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0047	183.6	34.00	283.2	89.50	22.04	15	-1.11	642.16	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0505	222.7	29.76	283.2	89.50	22.04	15	-0.68	642.16	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2192	265.0	0.0000	283.2	89.50	22.04	15	-0.20	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2369	266.7	3.290	283.2	89.50	22.04	15	-0.18	642.04	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0407	269.6	8.020	283.2	89.50	22.04	15	-0.15	642.99	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0227	276.5	19.00	283.2	89.50	22.04	15	-0.08	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0353	317.0	2.000	283.2	89.50	22.04	15	0.38	642.04	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0529	356.0	20.00	283.2	89.50	22.04	15	0.81	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0590	375.5	101.0	283.2	89.50	22.04	15	1.03	642.98	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0596	394.1	65.20	283.2	89.50	22.04	15	1.24	642.98	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2269	401.2	51.40	283.2	89.50	22.04	15	1.32	642.98	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2022	445.6	13.05	283.2	89.50	22.04	15	1.81	642.21	0
650	Zearalenone (ppb)	0242	110.5	23.00	340.7	107.1	22.51	43	-2.15	650.24	0
650	Zearalenone (ppb)	2129	136.0	2.000	340.7	107.1	22.51	43	-1.91	650.24	0
650	Zearalenone (ppb)	2268	140.0	40.00	340.7	107.1	22.51	43	-1.87	650.99	0
650	Zearalenone (ppb)	0656	149.2	0.0000	340.7	107.1	22.51	43	-1.79	650.06	0
650	Zearalenone (ppb)	2326	162.0	10.00	340.7	107.1	22.51	43	-1.67	650.23	0
650	Zearalenone (ppb)	0629	165.5	3.000	340.7	107.1	22.51	43	-1.64	650.05	0
650	Zearalenone (ppb)	2443	245.2	0.0000	340.7	107.1	22.51	43	-0.89	650.24	0
650	Zearalenone (ppb)	2113	248.5	49.00	340.7	107.1	22.51	43	-0.86	650.24	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT	Your	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests	Z Score	Method	
650	Zearalenone (ppb)	0066	263.2	41.29	340.7	107.1	22.51	43	-0.72	650.07	0
650	Zearalenone (ppb)	0689	269.0	36.00	340.7	107.1	22.51	43	-0.67	650.16	0
650	Zearalenone (ppb)	0098	285.0	30.00	340.7	107.1	22.51	43	-0.52	650.20	0
650	Zearalenone (ppb)	0870	288.4	4.154	340.7	107.1	22.51	43	-0.49	650.01	0
650	Zearalenone (ppb)	2033	290.0	0.0000	340.7	107.1	22.51	43	-0.47	650.24	0
650	Zearalenone (ppb)	0027	304.4	3.890	340.7	107.1	22.51	43	-0.34	650.01	0
650	Zearalenone (ppb)	0590	306.0	80.00	340.7	107.1	22.51	43	-0.32	650.06	0
650	Zearalenone (ppb)	0563	306.0	8.000	340.7	107.1	22.51	43	-0.32	650.20	0
650	Zearalenone (ppb)	0619	307.0	4.000	340.7	107.1	22.51	43	-0.31	650.24	0
650	Zearalenone (ppb)	2113	310.0	40.00	340.7	107.1	22.51	43	-0.29	650.01	0
650	Zearalenone (ppb)	0354	315.8	3.890	340.7	107.1	22.51	43	-0.23	650.20	0
650	Zearalenone (ppb)	0354	315.8	3.890	340.7	107.1	22.51	43	-0.23	650.21	0
650	Zearalenone (ppb)	0682	331.1	0.0000	340.7	107.1	22.51	43	-0.09	650.24	0
650	Zearalenone (ppb)	0964	348.3	43.50	340.7	107.1	22.51	43	0.07	650.24	0
650	Zearalenone (ppb)	2055	352.5	45.00	340.7	107.1	22.51	43	0.11	650.24	0
650	Zearalenone (ppb)	2369	362.7	37.64	340.7	107.1	22.51	43	0.21	650.20	0
650	Zearalenone (ppb)	2369	362.7	37.64	340.7	107.1	22.51	43	0.21	650.21	0
650	Zearalenone (ppb)	2302	369.5	13.00	340.7	107.1	22.51	43	0.27	650.99	0
650	Zearalenone (ppb)	2269	370.9	10.10	340.7	107.1	22.51	43	0.28	650.98	0
650	Zearalenone (ppb)	2140	371.0	18.00	340.7	107.1	22.51	43	0.28	650.24	0
650	Zearalenone (ppb)	0010	371.5	97.00	340.7	107.1	22.51	43	0.29	650.24	0
650	Zearalenone (ppb)	0148	373.6	25.78	340.7	107.1	22.51	43	0.31	650.24	0
650	Zearalenone (ppb)	0529	383.3	16.20	340.7	107.1	22.51	43	0.40	650.01	0
650	Zearalenone (ppb)	0353	392.0	26.00	340.7	107.1	22.51	43	0.48	650.23	0
650	Zearalenone (ppb)	0674	417.0	0.0000	340.7	107.1	22.51	43	0.71	650.20	0
650	Zearalenone (ppb)	2192	437.0	18.00	340.7	107.1	22.51	43	0.90	650.20	0
650	Zearalenone (ppb)	0505	461.1	5.490	340.7	107.1	22.51	43	1.12	650.16	0
650	Zearalenone (ppb)	0297	467.5	3.000	340.7	107.1	22.51	43	1.18	650.24	0
650	Zearalenone (ppb)	0047	469.7	5.700	340.7	107.1	22.51	43	1.20	650.16	0
650	Zearalenone (ppb)	0407	473.1	58.82	340.7	107.1	22.51	43	1.24	650.07	0
650	Zearalenone (ppb)	0035	473.1	12.41	340.7	107.1	22.51	43	1.24	650.07	0
650	Zearalenone (ppb)	2327	482.5	5.000	340.7	107.1	22.51	43	1.32	650.24	0
650	Zearalenone (ppb)	0354	495.4	5.220	340.7	107.1	22.51	43	1.44	650.04	0
650	Zearalenone (ppb)	0596	766.3	55.40	340.7	107.1	22.51	43	3.97	650.98	0
650	Zearalenone (ppb)	2022	784.2	12.76	340.7	107.1	22.51	43	4.14	650.04	0
650	Zearalenone (ppb)	0572	< 500		340.7	107.1	22.51	43		650.24	5

Note 1: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or between -2 and -3 (warning) and Green = Z <= 2 and >= -2 (OK at 95%). Horwitz SD's assigned based on Rob Mean only.

Note 2: Data Not Used Flags: 1=Rejected for duplicates too far apart, 2=Rejected as extreme outlier, 5=Reporting limit (<), 4=Zeros submitted as values and 3=Statistical problem. Flag 0 indicates data used in calculations.

Note 3: *Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). A zero range is not included in robust R-bar calculation.