



AAFCO
Association of American Feed Control Officials



**Mycotoxin Contaminants Scheme
Fish Feed**

**# Labs Reporting: 65
Analytes Reported 15
Issue Date : 11/30/2023**

Test Material Code # 202364

Analyte Proficiency Testing Report

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests			
600	Total Aflatoxin (ppb)	0656	5.870	0.0000	19.72	6.696	1.788	53	-2.07	600.12	0
600	Total Aflatoxin (ppb)	2192	11.70	0.6000	19.72	6.696	1.788	53	-1.20	600.20	0
600	Total Aflatoxin (ppb)	0723	11.82	0.0000	19.72	6.696	1.788	53	-1.18	600.20	0
600	Total Aflatoxin (ppb)	2268	13.00	0.0000	19.72	6.696	1.788	53	-1.00	600.14	0
600	Total Aflatoxin (ppb)	0629	13.38	0.1500	19.72	6.696	1.788	53	-0.95	600.20	0
600	Total Aflatoxin (ppb)	2113	14.25	1.100	19.72	6.696	1.788	53	-0.82	600.01	0
600	Total Aflatoxin (ppb)	2022	14.25	0.7000	19.72	6.696	1.788	53	-0.82	600.08	0
600	Total Aflatoxin (ppb)	2033	14.45	0.5000	19.72	6.696	1.788	53	-0.79	600.24	0
600	Total Aflatoxin (ppb)	0407	14.76	3.330	19.72	6.696	1.788	53	-0.74	600.13	0
600	Total Aflatoxin (ppb)	2456	14.77	1.687	19.72	6.696	1.788	53	-0.74	600.99	0
600	Total Aflatoxin (ppb)	0066	15.32	2.030	19.72	6.696	1.788	53	-0.66	600.13	0
600	Total Aflatoxin (ppb)	0682	16.40	0.0000	19.72	6.696	1.788	53	-0.50	600.24	0
600	Total Aflatoxin (ppb)	0619	16.50	2.000	19.72	6.696	1.788	53	-0.48	600.24	0
600	Total Aflatoxin (ppb)	0010	16.75	0.7000	19.72	6.696	1.788	53	-0.44	600.20	0
600	Total Aflatoxin (ppb)	0208	16.85	2.100	19.72	6.696	1.788	53	-0.43	600.01	0
600	Total Aflatoxin (ppb)	2199	17.50	11.00	19.72	6.696	1.788	53	-0.33	600.14	0
600	Total Aflatoxin (ppb)	2196	17.64	0.0000	19.72	6.696	1.788	53	-0.31	600.20	0
600	Total Aflatoxin (ppb)	2369	18.22	1.510	19.72	6.696	1.788	53	-0.22	600.14	0
600	Total Aflatoxin (ppb)	2369	18.22	1.510	19.72	6.696	1.788	53	-0.22	600.20	0
600	Total Aflatoxin (ppb)	2269	18.55	4.300	19.72	6.696	1.788	53	-0.17	600.98	0
600	Total Aflatoxin (ppb)	0674	18.55	0.0000	19.72	6.696	1.788	53	-0.17	600.20	0
600	Total Aflatoxin (ppb)	0354	19.07	5.770	19.72	6.696	1.788	53	-0.10	600.08	0
600	Total Aflatoxin (ppb)	2226	19.16	0.6620	19.72	6.696	1.788	53	-0.08	600.12	0
600	Total Aflatoxin (ppb)	2302	19.50	3.000	19.72	6.696	1.788	53	-0.03	600.99	0
600	Total Aflatoxin (ppb)	2096	19.70	5.600	19.72	6.696	1.788	53	0.00	600.01	0
600	Total Aflatoxin (ppb)	2129	19.75	0.9000	19.72	6.696	1.788	53	0.01	600.24	0
600	Total Aflatoxin (ppb)	0354	19.82	1.530	19.72	6.696	1.788	53	0.01	600.14	0
600	Total Aflatoxin (ppb)	0354	19.82	1.530	19.72	6.696	1.788	53	0.01	600.21	0
600	Total Aflatoxin (ppb)	0354	19.82	1.530	19.72	6.696	1.788	53	0.01	600.22	0
600	Total Aflatoxin (ppb)	0098	19.90	2.680	19.72	6.696	1.788	53	0.03	600.22	0
600	Total Aflatoxin (ppb)	0563	20.39	1.350	19.72	6.696	1.788	53	0.10	600.20	0
600	Total Aflatoxin (ppb)	0511	20.70	0.8000	19.72	6.696	1.788	53	0.15	600.01	0
600	Total Aflatoxin (ppb)	0001	20.90	1.400	19.72	6.696	1.788	53	0.18	600.20	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT	Your	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests	Z Score	Method	
600	Total Aflatoxin (ppb)	0353	20.98	0.8870	19.72	6.696	1.788	53	0.19	600.24	0
600	Total Aflatoxin (ppb)	0034	21.33	1.950	19.72	6.696	1.788	53	0.24	600.01	0
600	Total Aflatoxin (ppb)	0027	21.40	6.200	19.72	6.696	1.788	53	0.25	600.01	0
600	Total Aflatoxin (ppb)	0876	21.45	2.500	19.72	6.696	1.788	53	0.26	600.01	0
600	Total Aflatoxin (ppb)	0227	21.50	3.000	19.72	6.696	1.788	53	0.27	600.01	0
600	Total Aflatoxin (ppb)	2196	21.65	0.0000	19.72	6.696	1.788	53	0.29	600.08	0
600	Total Aflatoxin (ppb)	0918	21.75	1.010	19.72	6.696	1.788	53	0.30	600.01	0
600	Total Aflatoxin (ppb)	0004	22.00	0.0000	19.72	6.696	1.788	53	0.34	600.14	0
600	Total Aflatoxin (ppb)	0520	22.50	1.000	19.72	6.696	1.788	53	0.42	600.14	0
600	Total Aflatoxin (ppb)	0042	23.35	3.400	19.72	6.696	1.788	53	0.54	600.20	0
600	Total Aflatoxin (ppb)	0148	24.07	1.210	19.72	6.696	1.788	53	0.65	600.24	0
600	Total Aflatoxin (ppb)	0529	24.55	0.5000	19.72	6.696	1.788	53	0.72	600.01	0
600	Total Aflatoxin (ppb)	0554	24.75	0.3000	19.72	6.696	1.788	53	0.75	600.01	0
600	Total Aflatoxin (ppb)	0596	26.80	0.0000	19.72	6.696	1.788	53	1.06	600.98	0
600	Total Aflatoxin (ppb)	0035	27.11	1.460	19.72	6.696	1.788	53	1.10	600.21	0
600	Total Aflatoxin (ppb)	2345	28.00	2.000	19.72	6.696	1.788	53	1.24	600.14	0
600	Total Aflatoxin (ppb)	0297	29.98	3.380	19.72	6.696	1.788	53	1.53	600.24	0
600	Total Aflatoxin (ppb)	0590	31.60	3.200	19.72	6.696	1.788	53	1.77	600.12	0
600	Total Aflatoxin (ppb)	0015	31.70	0.0553	19.72	6.696	1.788	53	1.79	600.16	0
600	Total Aflatoxin (ppb)	2239	35.80	1.000	19.72	6.696	1.788	53	2.40	600.98	0
600	Total Aflatoxin (ppb)	0014	20.75	21.70	19.72	6.696	1.788	53	0.15	600.01	1
600	Total Aflatoxin (ppb)	2326	0.6500	1.300	19.72	6.696	1.788	53		600.23	4
601	Aflatoxin B1 (ppb)	2302	3.000	0.0000	17.96	6.117	1.010	34	-2.45	601.99	0
601	Aflatoxin B1 (ppb)	0689	10.05	0.5000	17.96	6.117	1.010	34	-1.29	601.16	0
601	Aflatoxin B1 (ppb)	0723	10.66	0.0000	17.96	6.117	1.010	34	-1.19	601.20	0
601	Aflatoxin B1 (ppb)	2192	10.92	0.6200	17.96	6.117	1.010	34	-1.15	601.20	0
601	Aflatoxin B1 (ppb)	0629	12.53	0.1700	17.96	6.117	1.010	34	-0.89	601.20	0
601	Aflatoxin B1 (ppb)	2055	12.75	0.5000	17.96	6.117	1.010	34	-0.85	601.23	0
601	Aflatoxin B1 (ppb)	2033	13.45	0.5000	17.96	6.117	1.010	34	-0.74	601.23	0
601	Aflatoxin B1 (ppb)	0028	13.49	0.5390	17.96	6.117	1.010	34	-0.73	601.23	0
601	Aflatoxin B1 (ppb)	0870	13.59	0.3750	17.96	6.117	1.010	34	-0.72	601.23	0
601	Aflatoxin B1 (ppb)	0010	15.35	0.5000	17.96	6.117	1.010	34	-0.43	601.20	0
601	Aflatoxin B1 (ppb)	0682	15.40	0.0000	17.96	6.117	1.010	34	-0.42	601.23	0
601	Aflatoxin B1 (ppb)	0619	15.85	0.7000	17.96	6.117	1.010	34	-0.35	601.23	0
601	Aflatoxin B1 (ppb)	2196	15.90	0.0000	17.96	6.117	1.010	34	-0.34	601.20	0
601	Aflatoxin B1 (ppb)	0918	16.35	0.3000	17.96	6.117	1.010	34	-0.26	601.23	0
601	Aflatoxin B1 (ppb)	2369	16.87	1.430	17.96	6.117	1.010	34	-0.18	601.20	0
601	Aflatoxin B1 (ppb)	2369	16.87	1.430	17.96	6.117	1.010	34	-0.18	601.24	0
601	Aflatoxin B1 (ppb)	0675	17.87	1.450	17.96	6.117	1.010	34	-0.02	601.20	0
601	Aflatoxin B1 (ppb)	0098	18.34	2.440	17.96	6.117	1.010	34	0.06	601.24	0
601	Aflatoxin B1 (ppb)	0354	18.93	1.100	17.96	6.117	1.010	34	0.16	601.20	0
601	Aflatoxin B1 (ppb)	0354	18.93	1.100	17.96	6.117	1.010	34	0.16	601.24	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT	Your	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests	Z Score	Method	
601	Aflatoxin B1 (ppb)	0353	19.58	1.020	17.96	6.117	1.010	34	0.26	601.23	0
601	Aflatoxin B1 (ppb)	0009	19.62	1.053	17.96	6.117	1.010	34	0.27	601.23	0
601	Aflatoxin B1 (ppb)	0878	19.65	0.1000	17.96	6.117	1.010	34	0.28	601.98	0
601	Aflatoxin B1 (ppb)	2269	19.67	3.760	17.96	6.117	1.010	34	0.28	601.98	0
601	Aflatoxin B1 (ppb)	0878	20.25	0.1000	17.96	6.117	1.010	34	0.37	601.21	0
601	Aflatoxin B1 (ppb)	0563	20.39	1.350	17.96	6.117	1.010	34	0.40	601.20	0
601	Aflatoxin B1 (ppb)	2140	21.00	2.000	17.96	6.117	1.010	34	0.50	601.23	0
601	Aflatoxin B1 (ppb)	2238	22.66	0.3450	17.96	6.117	1.010	34	0.77	601.98	0
601	Aflatoxin B1 (ppb)	0148	24.07	1.210	17.96	6.117	1.010	34	1.00	601.23	0
601	Aflatoxin B1 (ppb)	0035	24.88	1.730	17.96	6.117	1.010	34	1.13	601.21	0
601	Aflatoxin B1 (ppb)	0297	27.62	1.348	17.96	6.117	1.010	34	1.58	601.23	0
601	Aflatoxin B1 (ppb)	2327	30.45	3.300	17.96	6.117	1.010	34	2.04	601.23	0
601	Aflatoxin B1 (ppb)	2239	30.98	0.0500	17.96	6.117	1.010	34	2.13	601.98	0
601	Aflatoxin B1 (ppb)	0505	40.68	4.080	17.96	6.117	1.010	34	3.71	601.16	0
601	Aflatoxin B1 (ppb)	2326	0.6500	1.300	17.96	6.117	1.010	34		601.23	4
601	Aflatoxin B1 (ppb)	0572	< 10		17.96	6.117	1.010	34		601.23	5
602	Aflatoxin B2 (ppb)	2192	0.7750	0.0100	1.319	0.4821	0.1409	19	-1.13	602.20	0
602	Aflatoxin B2 (ppb)	0629	0.8500	0.0200	1.319	0.4821	0.1409	19	-0.97	602.20	0
602	Aflatoxin B2 (ppb)	0723	0.9700	0.0000	1.319	0.4821	0.1409	19	-0.72	602.20	0
602	Aflatoxin B2 (ppb)	0682	1.000	0.0000	1.319	0.4821	0.1409	19	-0.66	602.23	0
602	Aflatoxin B2 (ppb)	2033	1.000	0.0000	1.319	0.4821	0.1409	19	-0.66	602.23	0
602	Aflatoxin B2 (ppb)	0354	1.060	0.0800	1.319	0.4821	0.1409	19	-0.54	602.20	0
602	Aflatoxin B2 (ppb)	0354	1.060	0.0800	1.319	0.4821	0.1409	19	-0.54	602.24	0
602	Aflatoxin B2 (ppb)	0870	1.125	0.0500	1.319	0.4821	0.1409	19	-0.40	602.23	0
602	Aflatoxin B2 (ppb)	2369	1.350	0.0800	1.319	0.4821	0.1409	19	0.06	602.20	0
602	Aflatoxin B2 (ppb)	2369	1.350	0.0800	1.319	0.4821	0.1409	19	0.06	602.24	0
602	Aflatoxin B2 (ppb)	0675	1.360	0.1400	1.319	0.4821	0.1409	19	0.08	602.20	0
602	Aflatoxin B2 (ppb)	0010	1.400	0.2000	1.319	0.4821	0.1409	19	0.17	602.20	0
602	Aflatoxin B2 (ppb)	0353	1.406	0.1330	1.319	0.4821	0.1409	19	0.18	602.23	0
602	Aflatoxin B2 (ppb)	2327	1.475	0.3700	1.319	0.4821	0.1409	19	0.32	602.23	0
602	Aflatoxin B2 (ppb)	0098	1.560	0.2400	1.319	0.4821	0.1409	19	0.50	602.24	0
602	Aflatoxin B2 (ppb)	2196	1.750	0.0000	1.319	0.4821	0.1409	19	0.89	602.20	0
602	Aflatoxin B2 (ppb)	0918	1.750	0.1000	1.319	0.4821	0.1409	19	0.89	602.23	0
602	Aflatoxin B2 (ppb)	0035	2.235	0.2700	1.319	0.4821	0.1409	19	1.90	602.21	0
602	Aflatoxin B2 (ppb)	0297	2.350	2.100	1.319	0.4821	0.1409	19	2.14	602.23	0
602	Aflatoxin B2 (ppb)	2326	0.0000	0.0000	1.319	0.4821	0.1409	19		602.23	4
602	Aflatoxin B2 (ppb)	0619	0.6350	1.270	1.319	0.4821	0.1409	19		602.23	4
602	Aflatoxin B2 (ppb)	0009	< 0.63		1.319	0.4821	0.1409	19		602.23	5
602	Aflatoxin B2 (ppb)	0148	< 2.5		1.319	0.4821	0.1409	19		602.23	5
602	Aflatoxin B2 (ppb)	0563	< 5		1.319	0.4821	0.1409	19		602.20	5
602	Aflatoxin B2 (ppb)	2140	< 20		1.319	0.4821	0.1409	19		602.23	5
603	Aflatoxin G1 (ppb)	0505	2.130	0.3800				2		603.16	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests			
603	Aflatoxin G1 (ppb)	2302	16.50	3.000				2		603.99	0
603	Aflatoxin G1 (ppb)	0675	0.0000	0.0000				2		603.20	4
603	Aflatoxin G1 (ppb)	0619	0.0000	0.0000				2		603.23	4
603	Aflatoxin G1 (ppb)	2326	0.0000	0.0000				2		603.23	4
603	Aflatoxin G1 (ppb)	2327	< 0.2					2		603.23	5
603	Aflatoxin G1 (ppb)	0629	< 0.25					2		603.20	5
603	Aflatoxin G1 (ppb)	0682	< 0.5					2		603.23	5
603	Aflatoxin G1 (ppb)	0870	< 0.6					2		603.23	5
603	Aflatoxin G1 (ppb)	0009	< 0.69					2		603.23	5
603	Aflatoxin G1 (ppb)	0918	< 1					2		603.23	5
603	Aflatoxin G1 (ppb)	2033	< 1					2		603.23	5
603	Aflatoxin G1 (ppb)	0148	< 2.5					2		603.23	5
603	Aflatoxin G1 (ppb)	0098	< 15					2		603.24	5
603	Aflatoxin G1 (ppb)	2140	< 20					2		603.23	5
604	Aflatoxin G2 (ppb)	0675	0.0000	0.0000				0		604.20	4
604	Aflatoxin G2 (ppb)	0619	0.0000	0.0000				0		604.23	4
604	Aflatoxin G2 (ppb)	2326	0.0000	0.0000				0		604.23	4
604	Aflatoxin G2 (ppb)	2327	< 0.2					0		604.23	5
604	Aflatoxin G2 (ppb)	0629	< 0.25					0		604.20	5
604	Aflatoxin G2 (ppb)	0682	< 0.5					0		604.23	5
604	Aflatoxin G2 (ppb)	0009	< 0.69					0		604.23	5
604	Aflatoxin G2 (ppb)	0870	< 1					0		604.23	5
604	Aflatoxin G2 (ppb)	0918	< 1					0		604.23	5
604	Aflatoxin G2 (ppb)	2033	< 1					0		604.23	5
604	Aflatoxin G2 (ppb)	0148	< 3					0		604.23	5
604	Aflatoxin G2 (ppb)	0098	< 5					0		604.24	5
604	Aflatoxin G2 (ppb)	2140	< 20					0		604.23	5
610	Deoxynivalenol (ppb)	0297	1.312	0.0002	1,452	439.0	103.0	57	-3.31	610.23	0
610	Deoxynivalenol (ppb)	0689	550.5	237.0	1,452	439.0	103.0	57	-2.05	610.16	0
610	Deoxynivalenol (ppb)	0066	685.0	30.00	1,452	439.0	103.0	57	-1.75	610.08	0
610	Deoxynivalenol (ppb)	2129	748.0	28.00	1,452	439.0	103.0	57	-1.60	610.23	0
610	Deoxynivalenol (ppb)	2302	826.5	37.00	1,452	439.0	103.0	57	-1.43	610.99	0
610	Deoxynivalenol (ppb)	0028	879.4	19.90	1,452	439.0	103.0	57	-1.30	610.23	0
610	Deoxynivalenol (ppb)	0572	880.0	120.0	1,452	439.0	103.0	57	-1.30	610.23	0
610	Deoxynivalenol (ppb)	2268	903.0	64.00	1,452	439.0	103.0	57	-1.25	610.22	0
610	Deoxynivalenol (ppb)	2456	918.4	112.7	1,452	439.0	103.0	57	-1.22	610.99	0
610	Deoxynivalenol (ppb)	2055	925.5	22.30	1,452	439.0	103.0	57	-1.20	610.23	0
610	Deoxynivalenol (ppb)	2326	949.0	88.00	1,452	439.0	103.0	57	-1.15	610.23	0
610	Deoxynivalenol (ppb)	0964	1,004	117.5	1,452	439.0	103.0	57	-1.02	610.23	0
610	Deoxynivalenol (ppb)	0015	1,085	63.50	1,452	439.0	103.0	57	-0.84	610.16	0
610	Deoxynivalenol (ppb)	0619	1,090	25.00	1,452	439.0	103.0	57	-0.83	610.23	0
610	Deoxynivalenol (ppb)	0656	1,143	0.0000	1,452	439.0	103.0	57	-0.70	610.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT	Your	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests	Z Score	Method	
610	Deoxynivalenol (ppb)	2033	1,200	0.0000	1,452	439.0	103.0	57	-0.57	610.23	0
610	Deoxynivalenol (ppb)	0010	1,227	58.00	1,452	439.0	103.0	57	-0.51	610.23	0
610	Deoxynivalenol (ppb)	2369	1,239	0.0000	1,452	439.0	103.0	57	-0.49	610.20	0
610	Deoxynivalenol (ppb)	0596	1,300	0.0000	1,452	439.0	103.0	57	-0.35	610.98	0
610	Deoxynivalenol (ppb)	2140	1,325	93.00	1,452	439.0	103.0	57	-0.29	610.23	0
610	Deoxynivalenol (ppb)	0870	1,356	44.50	1,452	439.0	103.0	57	-0.22	610.99	0
610	Deoxynivalenol (ppb)	2269	1,379	17.00	1,452	439.0	103.0	57	-0.17	610.98	0
610	Deoxynivalenol (ppb)	0148	1,386	81.62	1,452	439.0	103.0	57	-0.15	610.23	0
610	Deoxynivalenol (ppb)	0505	1,393	92.41	1,452	439.0	103.0	57	-0.13	610.16	0
610	Deoxynivalenol (ppb)	0098	1,400	200.0	1,452	439.0	103.0	57	-0.12	610.20	0
610	Deoxynivalenol (ppb)	2196	1,430	0.0000	1,452	439.0	103.0	57	-0.05	610.20	0
610	Deoxynivalenol (ppb)	0009	1,437	44.29	1,452	439.0	103.0	57	-0.03	610.23	0
610	Deoxynivalenol (ppb)	2327	1,469	25.00	1,452	439.0	103.0	57	0.04	610.23	0
610	Deoxynivalenol (ppb)	0918	1,489	130.0	1,452	439.0	103.0	57	0.08	610.01	0
610	Deoxynivalenol (ppb)	0674	1,489	0.0000	1,452	439.0	103.0	57	0.08	610.20	0
610	Deoxynivalenol (ppb)	0682	1,490	0.0000	1,452	439.0	103.0	57	0.09	610.23	0
610	Deoxynivalenol (ppb)	2239	1,508	15.00	1,452	439.0	103.0	57	0.13	610.98	0
610	Deoxynivalenol (ppb)	0407	1,532	157.0	1,452	439.0	103.0	57	0.18	610.07	0
610	Deoxynivalenol (ppb)	0529	1,550	500.0	1,452	439.0	103.0	57	0.22	610.01	0
610	Deoxynivalenol (ppb)	2196	1,571	0.0000	1,452	439.0	103.0	57	0.27	610.05	0
610	Deoxynivalenol (ppb)	0870	1,590	60.00	1,452	439.0	103.0	57	0.31	610.23	0
610	Deoxynivalenol (ppb)	0878	1,620	405.0	1,452	439.0	103.0	57	0.38	610.98	0
610	Deoxynivalenol (ppb)	0629	1,630	55.00	1,452	439.0	103.0	57	0.40	610.07	0
610	Deoxynivalenol (ppb)	0353	1,637	22.93	1,452	439.0	103.0	57	0.42	610.23	0
610	Deoxynivalenol (ppb)	0034	1,650	0.0000	1,452	439.0	103.0	57	0.45	610.01	0
610	Deoxynivalenol (ppb)	0027	1,670	280.0	1,452	439.0	103.0	57	0.50	610.01	0
610	Deoxynivalenol (ppb)	0590	1,708	235.0	1,452	439.0	103.0	57	0.58	610.07	0
610	Deoxynivalenol (ppb)	0918	1,710	76.04	1,452	439.0	103.0	57	0.59	610.23	0
610	Deoxynivalenol (ppb)	0878	1,745	18.00	1,452	439.0	103.0	57	0.67	610.20	0
610	Deoxynivalenol (ppb)	0554	1,800	0.0000	1,452	439.0	103.0	57	0.79	610.01	0
610	Deoxynivalenol (ppb)	2113	1,800	200.0	1,452	439.0	103.0	57	0.79	610.01	0
610	Deoxynivalenol (ppb)	0208	1,875	190.0	1,452	439.0	103.0	57	0.96	610.01	0
610	Deoxynivalenol (ppb)	0014	1,950	700.0	1,452	439.0	103.0	57	1.13	610.01	0
610	Deoxynivalenol (ppb)	2022	1,970	40.00	1,452	439.0	103.0	57	1.18	610.05	0
610	Deoxynivalenol (ppb)	0227	2,000	400.0	1,452	439.0	103.0	57	1.25	610.01	0
610	Deoxynivalenol (ppb)	0354	2,020	83.06	1,452	439.0	103.0	57	1.29	610.05	0
610	Deoxynivalenol (ppb)	2238	2,044	187.6	1,452	439.0	103.0	57	1.35	610.98	0
610	Deoxynivalenol (ppb)	0563	2,067	31.00	1,452	439.0	103.0	57	1.40	610.25	0
610	Deoxynivalenol (ppb)	0511	2,100	0.0000	1,452	439.0	103.0	57	1.48	610.01	0
610	Deoxynivalenol (ppb)	0035	2,103	11.00	1,452	439.0	103.0	57	1.48	610.08	0
610	Deoxynivalenol (ppb)	0042	2,200	200.0	1,452	439.0	103.0	57	1.70	610.01	0
610	Deoxynivalenol (ppb)	2411	4,625	190.0	1,452	439.0	103.0	57	7.23	610.20	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT	Your	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests	Z Score	Method	
610	Deoxynivalenol (ppb)	2345	3,330	2,900	1,452	439.0	103.0	57	4.28	610.09	1
620	Total Fumonisin (ppb)	2369	1.385	0.0500	1,683	506.7	176.4	36	-3.32	620.11	0
620	Total Fumonisin (ppb)	0619	1.808	0.1210	1,683	506.7	176.4	36	-3.32	620.23	0
620	Total Fumonisin (ppb)	2387	1.950	0.1000	1,683	506.7	176.4	36	-3.32	620.98	0
620	Total Fumonisin (ppb)	0689	539.0	60.00	1,683	506.7	176.4	36	-2.26	620.16	0
620	Total Fumonisin (ppb)	2302	1,037	86.00	1,683	506.7	176.4	36	-1.27	620.99	0
620	Total Fumonisin (ppb)	0015	1,075	35.43	1,683	506.7	176.4	36	-1.20	620.16	0
620	Total Fumonisin (ppb)	0656	1,175	1,800	1,683	506.7	176.4	36	-1.00	620.09	0
620	Total Fumonisin (ppb)	0505	1,255	77.31	1,683	506.7	176.4	36	-0.84	620.16	0
620	Total Fumonisin (ppb)	0674	1,310	0.0000	1,683	506.7	176.4	36	-0.74	620.07	0
620	Total Fumonisin (ppb)	0407	1,355	130.0	1,683	506.7	176.4	36	-0.65	620.09	0
620	Total Fumonisin (ppb)	2456	1,414	195.2	1,683	506.7	176.4	36	-0.53	620.99	0
620	Total Fumonisin (ppb)	0035	1,485	30.00	1,683	506.7	176.4	36	-0.39	620.09	0
620	Total Fumonisin (ppb)	0529	1,550	300.0	1,683	506.7	176.4	36	-0.26	620.02	0
620	Total Fumonisin (ppb)	0034	1,590	280.0	1,683	506.7	176.4	36	-0.18	620.01	0
620	Total Fumonisin (ppb)	0014	1,620	1,140	1,683	506.7	176.4	36	-0.12	620.01	0
620	Total Fumonisin (ppb)	2239	1,629	36.00	1,683	506.7	176.4	36	-0.11	620.98	0
620	Total Fumonisin (ppb)	0878	1,650	51.00	1,683	506.7	176.4	36	-0.07	620.98	0
620	Total Fumonisin (ppb)	0918	1,655	472.0	1,683	506.7	176.4	36	-0.05	620.01	0
620	Total Fumonisin (ppb)	0682	1,740	0.0000	1,683	506.7	176.4	36	0.11	620.23	0
620	Total Fumonisin (ppb)	2268	1,750	100.0	1,683	506.7	176.4	36	0.13	620.11	0
620	Total Fumonisin (ppb)	0098	1,800	200.0	1,683	506.7	176.4	36	0.23	620.11	0
620	Total Fumonisin (ppb)	0010	1,855	182.0	1,683	506.7	176.4	36	0.34	620.23	0
620	Total Fumonisin (ppb)	0511	1,900	400.0	1,683	506.7	176.4	36	0.43	620.01	0
620	Total Fumonisin (ppb)	2033	1,904	21.00	1,683	506.7	176.4	36	0.44	620.23	0
620	Total Fumonisin (ppb)	0042	1,950	300.0	1,683	506.7	176.4	36	0.53	620.01	0
620	Total Fumonisin (ppb)	2113	1,950	100.0	1,683	506.7	176.4	36	0.53	620.01	0
620	Total Fumonisin (ppb)	0227	2,000	0.0000	1,683	506.7	176.4	36	0.63	620.01	0
620	Total Fumonisin (ppb)	0027	2,065	290.0	1,683	506.7	176.4	36	0.75	620.01	0
620	Total Fumonisin (ppb)	0353	2,089	31.14	1,683	506.7	176.4	36	0.80	620.23	0
620	Total Fumonisin (ppb)	2129	2,094	47.00	1,683	506.7	176.4	36	0.81	620.23	0
620	Total Fumonisin (ppb)	0520	2,100	200.0	1,683	506.7	176.4	36	0.82	620.12	0
620	Total Fumonisin (ppb)	0208	2,350	500.0	1,683	506.7	176.4	36	1.32	620.98	0
620	Total Fumonisin (ppb)	2022	2,380	80.00	1,683	506.7	176.4	36	1.38	620.07	0
620	Total Fumonisin (ppb)	0596	2,400	400.0	1,683	506.7	176.4	36	1.42	620.98	0
620	Total Fumonisin (ppb)	0066	2,665	270.0	1,683	506.7	176.4	36	1.94	620.09	0
620	Total Fumonisin (ppb)	0554	2,700	0.0000	1,683	506.7	176.4	36	2.01	620.01	0
621	Fumonsin B1 (ppb)	2369	1.095	0.0300	1,381	418.0	57.05	18	-3.30	621.21	0
621	Fumonsin B1 (ppb)	0572	555.0	10.00	1,381	418.0	57.05	18	-1.98	621.23	0
621	Fumonsin B1 (ppb)	2327	676.5	13.00	1,381	418.0	57.05	18	-1.68	621.23	0
621	Fumonsin B1 (ppb)	2326	1,010	324.0	1,381	418.0	57.05	18	-0.89	621.23	0
621	Fumonsin B1 (ppb)	2196	1,169	0.0000	1,381	418.0	57.05	18	-0.51	621.20	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT	Your	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests	Z Score	Method	
621	Fumonsin B1 (ppb)	0148	1,327	30.92	1,381	418.0	57.05	18	-0.13	621.23	0
621	Fumonsin B1 (ppb)	0619	1,394	74.00	1,381	418.0	57.05	18	0.03	621.23	0
621	Fumonsin B1 (ppb)	0010	1,407	82.00	1,381	418.0	57.05	18	0.06	621.23	0
621	Fumonsin B1 (ppb)	2140	1,415	63.00	1,381	418.0	57.05	18	0.08	621.23	0
621	Fumonsin B1 (ppb)	0682	1,430	0.0000	1,381	418.0	57.05	18	0.12	621.23	0
621	Fumonsin B1 (ppb)	0098	1,450	100.0	1,381	418.0	57.05	18	0.17	621.21	0
621	Fumonsin B1 (ppb)	0009	1,479	58.70	1,381	418.0	57.05	18	0.23	621.23	0
621	Fumonsin B1 (ppb)	0353	1,534	13.97	1,381	418.0	57.05	18	0.37	621.23	0
621	Fumonsin B1 (ppb)	2238	1,567	0.0000	1,381	418.0	57.05	18	0.45	621.98	0
621	Fumonsin B1 (ppb)	0870	1,590	20.00	1,381	418.0	57.05	18	0.50	621.23	0
621	Fumonsin B1 (ppb)	0918	1,883	91.85	1,381	418.0	57.05	18	1.20	621.23	0
621	Fumonsin B1 (ppb)	2055	2,134	34.90	1,381	418.0	57.05	18	1.80	621.23	0
621	Fumonsin B1 (ppb)	0563	2,408	275.0	1,381	418.0	57.05	18	2.46	621.20	0
621	Fumonsin B1 (ppb)	0964	9,359	549.4	1,381	418.0	57.05	18	19.09	621.23	2
622	Fumonisin B2 (ppb)	2369	0.2900	0.0400	410.7	128.5	45.76	16	-3.19	622.21	0
622	Fumonisin B2 (ppb)	2326	267.0	134.0	410.7	128.5	45.76	16	-1.12	622.23	0
622	Fumonisin B2 (ppb)	0010	284.5	255.0	410.7	128.5	45.76	16	-0.98	622.23	0
622	Fumonisin B2 (ppb)	0682	310.0	0.0000	410.7	128.5	45.76	16	-0.78	622.23	0
622	Fumonisin B2 (ppb)	0098	350.0	100.0	410.7	128.5	45.76	16	-0.47	622.21	0
622	Fumonisin B2 (ppb)	2140	355.0	2.000	410.7	128.5	45.76	16	-0.43	622.23	0
622	Fumonisin B2 (ppb)	2196	382.1	0.0000	410.7	128.5	45.76	16	-0.22	622.20	0
622	Fumonisin B2 (ppb)	0870	410.0	20.00	410.7	128.5	45.76	16	-0.01	622.23	0
622	Fumonisin B2 (ppb)	0619	414.0	48.00	410.7	128.5	45.76	16	0.03	622.23	0
622	Fumonisin B2 (ppb)	2055	416.8	9.300	410.7	128.5	45.76	16	0.05	622.23	0
622	Fumonisin B2 (ppb)	0148	430.3	3.170	410.7	128.5	45.76	16	0.15	622.23	0
622	Fumonisin B2 (ppb)	0353	467.3	3.870	410.7	128.5	45.76	16	0.44	622.23	0
622	Fumonisin B2 (ppb)	0009	471.1	6.990	410.7	128.5	45.76	16	0.47	622.23	0
622	Fumonisin B2 (ppb)	2327	578.0	6.000	410.7	128.5	45.76	16	1.30	622.23	0
622	Fumonisin B2 (ppb)	0918	770.1	19.13	410.7	128.5	45.76	16	2.80	622.23	0
622	Fumonisin B2 (ppb)	0964	3,235	152.1	410.7	128.5	45.76	16	21.99	622.23	0
623	Fumonisin B3 (ppb)	0353	82.16	0.6320	177.7	56.87	5.006	7	-1.68	623.23	0
623	Fumonisin B3 (ppb)	2326	110.5	43.00	177.7	56.87	5.006	7	-1.18	623.23	0
623	Fumonisin B3 (ppb)	0010	163.5	9.000	177.7	56.87	5.006	7	-0.25	623.23	0
623	Fumonisin B3 (ppb)	0009	195.6	1.560	177.7	56.87	5.006	7	0.31	623.23	0
623	Fumonisin B3 (ppb)	0870	200.0	0.0000	177.7	56.87	5.006	7	0.39	623.23	0
623	Fumonisin B3 (ppb)	2327	200.5	3.000	177.7	56.87	5.006	7	0.40	623.23	0
623	Fumonisin B3 (ppb)	0918	291.7	1.530	177.7	56.87	5.006	7	2.00	623.23	0
630	Ochratoxin A (ppb)	2456	2.831	0.3290	9.960	3.446	1.378	39	-2.07	630.99	0
630	Ochratoxin A (ppb)	0918	3.285	0.1900	9.960	3.446	1.378	39	-1.94	630.01	0
630	Ochratoxin A (ppb)	0407	4.015	1.590	9.960	3.446	1.378	39	-1.73	630.04	0
630	Ochratoxin A (ppb)	0596	4.160	3.520	9.960	3.446	1.378	39	-1.68	630.98	0
630	Ochratoxin A (ppb)	2022	4.245	0.2700	9.960	3.446	1.378	39	-1.66	630.03	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT	Your	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests	Z Score	Method	
630	Ochratoxin A (ppb)	0689	4.350	0.3000	9.960	3.446	1.378	39	-1.63	630.16	0
630	Ochratoxin A (ppb)	2326	4.500	3.000	9.960	3.446	1.378	39	-1.58	630.23	0
630	Ochratoxin A (ppb)	0870	5.420	0.8800	9.960	3.446	1.378	39	-1.32	630.23	0
630	Ochratoxin A (ppb)	2302	5.500	1.000	9.960	3.446	1.378	39	-1.29	630.99	0
630	Ochratoxin A (ppb)	2226	5.957	0.5790	9.960	3.446	1.378	39	-1.16	630.05	0
630	Ochratoxin A (ppb)	0723	6.780	0.0000	9.960	3.446	1.378	39	-0.92	630.20	0
630	Ochratoxin A (ppb)	0675	7.230	1.320	9.960	3.446	1.378	39	-0.79	630.20	0
630	Ochratoxin A (ppb)	0353	7.506	0.1470	9.960	3.446	1.378	39	-0.71	630.23	0
630	Ochratoxin A (ppb)	2033	8.100	0.2000	9.960	3.446	1.378	39	-0.54	630.23	0
630	Ochratoxin A (ppb)	0354	8.525	0.3100	9.960	3.446	1.378	39	-0.42	630.06	0
630	Ochratoxin A (ppb)	0354	8.525	0.3100	9.960	3.446	1.378	39	-0.42	630.20	0
630	Ochratoxin A (ppb)	0682	8.900	0.0000	9.960	3.446	1.378	39	-0.31	630.23	0
630	Ochratoxin A (ppb)	2369	9.430	1.660	9.960	3.446	1.378	39	-0.15	630.06	0
630	Ochratoxin A (ppb)	2369	9.430	1.660	9.960	3.446	1.378	39	-0.15	630.20	0
630	Ochratoxin A (ppb)	0878	9.450	0.1000	9.960	3.446	1.378	39	-0.15	630.98	0
630	Ochratoxin A (ppb)	0563	9.945	0.6100	9.960	3.446	1.378	39	0.00	630.20	0
630	Ochratoxin A (ppb)	0629	10.90	1.000	9.960	3.446	1.378	39	0.27	630.20	0
630	Ochratoxin A (ppb)	0520	11.00	0.0000	9.960	3.446	1.378	39	0.30	630.06	0
630	Ochratoxin A (ppb)	0010	11.00	4.000	9.960	3.446	1.378	39	0.30	630.23	0
630	Ochratoxin A (ppb)	0505	11.28	0.5700	9.960	3.446	1.378	39	0.38	630.16	0
630	Ochratoxin A (ppb)	0354	11.48	2.620	9.960	3.446	1.378	39	0.44	630.03	0
630	Ochratoxin A (ppb)	2239	12.05	1.300	9.960	3.446	1.378	39	0.61	630.98	0
630	Ochratoxin A (ppb)	2196	12.23	0.0000	9.960	3.446	1.378	39	0.66	630.20	0
630	Ochratoxin A (ppb)	0918	12.85	2.100	9.960	3.446	1.378	39	0.84	630.23	0
630	Ochratoxin A (ppb)	0656	13.35	0.0000	9.960	3.446	1.378	39	0.98	630.05	0
630	Ochratoxin A (ppb)	0066	13.74	1.730	9.960	3.446	1.378	39	1.10	630.05	0
630	Ochratoxin A (ppb)	2327	14.10	0.8000	9.960	3.446	1.378	39	1.20	630.23	0
630	Ochratoxin A (ppb)	0619	14.95	5.900	9.960	3.446	1.378	39	1.45	630.23	0
630	Ochratoxin A (ppb)	0674	15.55	0.0000	9.960	3.446	1.378	39	1.62	630.20	0
630	Ochratoxin A (ppb)	0148	16.14	1.880	9.960	3.446	1.378	39	1.79	630.23	0
630	Ochratoxin A (ppb)	2140	17.50	5.000	9.960	3.446	1.378	39	2.19	630.23	0
630	Ochratoxin A (ppb)	2238	21.43	1.921	9.960	3.446	1.378	39	3.33	630.98	0
630	Ochratoxin A (ppb)	2129	25.75	1.100	9.960	3.446	1.378	39	4.58	630.23	0
630	Ochratoxin A (ppb)	0529	118.2	20.60	9.960	3.446	1.378	39	31.41	630.01	0
630	Ochratoxin A (ppb)	0964	1,743	170.8	9.960	3.446	1.378	39	>100	630.23	2
630	Ochratoxin A (ppb)	0227	< 2		9.960	3.446	1.378	39		630.01	5
630	Ochratoxin A (ppb)	0098	< 10		9.960	3.446	1.378	39		630.20	5
630	Ochratoxin A (ppb)	0572	< 20		9.960	3.446	1.378	39		630.23	5
640	T-2 Toxin (ppb)	2326	34.50	23.00	95.74	31.15	6.723	25	-1.97	640.23	0
640	T-2 Toxin (ppb)	2369	54.59	2.410	95.74	31.15	6.723	25	-1.32	640.23	0
640	T-2 Toxin (ppb)	2129	67.65	1.500	95.74	31.15	6.723	25	-0.90	640.23	0
640	T-2 Toxin (ppb)	2033	70.00	0.0000	95.74	31.15	6.723	25	-0.83	640.23	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT	Your	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests	Z Score	Method	
640	T-2 Toxin (ppb)	0682	73.60	0.0000	95.74	31.15	6.723	25	-0.71	640.23	0
640	T-2 Toxin (ppb)	2055	79.45	11.70	95.74	31.15	6.723	25	-0.52	640.23	0
640	T-2 Toxin (ppb)	0297	84.19	3.910	95.74	31.15	6.723	25	-0.37	640.23	0
640	T-2 Toxin (ppb)	2140	88.50	9.000	95.74	31.15	6.723	25	-0.23	640.23	0
640	T-2 Toxin (ppb)	0674	89.50	0.0000	95.74	31.15	6.723	25	-0.20	640.03	0
640	T-2 Toxin (ppb)	2239	90.82	1.640	95.74	31.15	6.723	25	-0.16	640.98	0
640	T-2 Toxin (ppb)	2196	91.06	0.0000	95.74	31.15	6.723	25	-0.15	640.20	0
640	T-2 Toxin (ppb)	0870	92.44	8.800	95.74	31.15	6.723	25	-0.11	640.23	0
640	T-2 Toxin (ppb)	0619	93.80	9.400	95.74	31.15	6.723	25	-0.06	640.23	0
640	T-2 Toxin (ppb)	0148	98.84	12.51	95.74	31.15	6.723	25	0.10	640.23	0
640	T-2 Toxin (ppb)	0878	99.60	3.600	95.74	31.15	6.723	25	0.12	640.98	0
640	T-2 Toxin (ppb)	0563	101.5	3.000	95.74	31.15	6.723	25	0.18	640.25	0
640	T-2 Toxin (ppb)	0407	102.3	5.870	95.74	31.15	6.723	25	0.21	640.05	0
640	T-2 Toxin (ppb)	2327	103.5	3.000	95.74	31.15	6.723	25	0.25	640.23	0
640	T-2 Toxin (ppb)	0918	106.2	6.510	95.74	31.15	6.723	25	0.34	640.23	0
640	T-2 Toxin (ppb)	0353	109.2	0.8680	95.74	31.15	6.723	25	0.43	640.23	0
640	T-2 Toxin (ppb)	0015	113.5	3.447	95.74	31.15	6.723	25	0.57	640.16	0
640	T-2 Toxin (ppb)	0656	124.0	0.0000	95.74	31.15	6.723	25	0.91	640.05	0
640	T-2 Toxin (ppb)	2238	154.5	1.000	95.74	31.15	6.723	25	1.89	640.98	0
640	T-2 Toxin (ppb)	0066	156.7	24.47	95.74	31.15	6.723	25	1.96	640.03	0
640	T-2 Toxin (ppb)	2269	194.9	24.80	95.74	31.15	6.723	25	3.18	640.98	0
640	T-2 Toxin (ppb)	0964	3,288	122.0	95.74	31.15	6.723	25	>100	640.23	2
640	T-2 Toxin (ppb)	0572	< 100		95.74	31.15	6.723	25		640.23	5
641	HT-2 Toxin (ppb)	2326	61.00	58.00	118.7	38.41	16.55	12	-1.50	641.23	0
641	HT-2 Toxin (ppb)	0297	71.29	6.728	118.7	38.41	16.55	12	-1.24	641.23	0
641	HT-2 Toxin (ppb)	2369	81.93	4.460	118.7	38.41	16.55	12	-0.96	641.23	0
641	HT-2 Toxin (ppb)	2033	92.55	8.300	118.7	38.41	16.55	12	-0.68	641.23	0
641	HT-2 Toxin (ppb)	0870	107.0	0.9800	118.7	38.41	16.55	12	-0.31	641.23	0
641	HT-2 Toxin (ppb)	0619	112.4	27.20	118.7	38.41	16.55	12	-0.16	641.23	0
641	HT-2 Toxin (ppb)	2140	113.5	33.00	118.7	38.41	16.55	12	-0.14	641.23	0
641	HT-2 Toxin (ppb)	2327	125.5	13.00	118.7	38.41	16.55	12	0.18	641.23	0
641	HT-2 Toxin (ppb)	0148	145.5	23.00	118.7	38.41	16.55	12	0.70	641.23	0
641	HT-2 Toxin (ppb)	2055	154.8	28.20	118.7	38.41	16.55	12	0.94	641.23	0
641	HT-2 Toxin (ppb)	0918	172.9	9.960	118.7	38.41	16.55	12	1.41	641.23	0
641	HT-2 Toxin (ppb)	0353	230.3	5.741	118.7	38.41	16.55	12	2.90	641.23	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0689	66.50	7.000	241.0	76.49	17.71	15	-2.28	642.16	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2302	97.00	14.00	241.0	76.49	17.71	15	-1.88	642.99	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0227	107.0	22.00	241.0	76.49	17.71	15	-1.75	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2369	136.5	2.050	241.0	76.49	17.71	15	-1.37	642.04	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0297	155.5	2.818	241.0	76.49	17.71	15	-1.12	642.04	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0505	171.8	13.32	241.0	76.49	17.71	15	-0.91	642.16	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0629	236.0	0.0000	241.0	76.49	17.71	15	-0.07	642.98	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT	Your	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests	Z Score	Method	
642	T-2 Toxin + HT-2 Toxin (ppb)	2269	269.0	34.20	241.0	76.49	17.71	15	0.37	642.98	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2022	280.9	5.540	241.0	76.49	17.71	15	0.52	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0529	292.5	26.70	241.0	76.49	17.71	15	0.67	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0353	326.4	14.66	241.0	76.49	17.71	15	1.12	642.04	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0918	331.8	61.91	241.0	76.49	17.71	15	1.19	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0590	353.0	0.0000	241.0	76.49	17.71	15	1.46	642.98	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0407	359.8	32.34	241.0	76.49	17.71	15	1.55	642.99	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0596	440.7	17.12	241.0	76.49	17.71	15	2.61	642.98	0
650	Zearalenone (ppb)	0656	121.4	0.0000	271.8	85.99	20.70	50	-1.75	650.06	0
650	Zearalenone (ppb)	2326	128.0	72.00	271.8	85.99	20.70	50	-1.67	650.23	0
650	Zearalenone (ppb)	2129	156.0	4.000	271.8	85.99	20.70	50	-1.35	650.24	0
650	Zearalenone (ppb)	0354	179.7	30.06	271.8	85.99	20.70	50	-1.07	650.04	0
650	Zearalenone (ppb)	0689	181.0	46.00	271.8	85.99	20.70	50	-1.06	650.16	0
650	Zearalenone (ppb)	2196	181.4	0.0000	271.8	85.99	20.70	50	-1.05	650.20	0
650	Zearalenone (ppb)	2302	184.5	35.00	271.8	85.99	20.70	50	-1.02	650.99	0
650	Zearalenone (ppb)	0629	197.0	26.00	271.8	85.99	20.70	50	-0.87	650.05	0
650	Zearalenone (ppb)	2196	203.8	0.0000	271.8	85.99	20.70	50	-0.79	650.04	0
650	Zearalenone (ppb)	0066	212.3	30.90	271.8	85.99	20.70	50	-0.69	650.07	0
650	Zearalenone (ppb)	2369	227.8	2.670	271.8	85.99	20.70	50	-0.51	650.20	0
650	Zearalenone (ppb)	2369	227.8	2.670	271.8	85.99	20.70	50	-0.51	650.21	0
650	Zearalenone (ppb)	0619	233.5	15.00	271.8	85.99	20.70	50	-0.45	650.24	0
650	Zearalenone (ppb)	0407	234.7	101.1	271.8	85.99	20.70	50	-0.43	650.07	0
650	Zearalenone (ppb)	0682	234.7	0.0000	271.8	85.99	20.70	50	-0.43	650.24	0
650	Zearalenone (ppb)	0918	235.3	92.40	271.8	85.99	20.70	50	-0.43	650.01	0
650	Zearalenone (ppb)	2226	237.6	30.26	271.8	85.99	20.70	50	-0.40	650.06	0
650	Zearalenone (ppb)	0674	239.0	0.0000	271.8	85.99	20.70	50	-0.38	650.20	0
650	Zearalenone (ppb)	2022	240.7	3.430	271.8	85.99	20.70	50	-0.36	650.04	0
650	Zearalenone (ppb)	2033	250.0	20.00	271.8	85.99	20.70	50	-0.25	650.24	0
650	Zearalenone (ppb)	0870	254.1	1.518	271.8	85.99	20.70	50	-0.21	650.01	0
650	Zearalenone (ppb)	0027	256.8	1.800	271.8	85.99	20.70	50	-0.17	650.01	0
650	Zearalenone (ppb)	2055	259.5	19.10	271.8	85.99	20.70	50	-0.14	650.24	0
650	Zearalenone (ppb)	2269	261.8	42.39	271.8	85.99	20.70	50	-0.12	650.98	0
650	Zearalenone (ppb)	0034	272.2	24.70	271.8	85.99	20.70	50	0.00	650.01	0
650	Zearalenone (ppb)	0098	275.0	50.00	271.8	85.99	20.70	50	0.04	650.20	0
650	Zearalenone (ppb)	0520	275.0	10.00	271.8	85.99	20.70	50	0.04	650.99	0
650	Zearalenone (ppb)	0354	277.7	19.61	271.8	85.99	20.70	50	0.07	650.20	0
650	Zearalenone (ppb)	0354	277.7	19.61	271.8	85.99	20.70	50	0.07	650.21	0
650	Zearalenone (ppb)	0878	289.6	30.90	271.8	85.99	20.70	50	0.21	650.98	0
650	Zearalenone (ppb)	0629	293.0	16.00	271.8	85.99	20.70	50	0.25	650.20	0
650	Zearalenone (ppb)	0353	293.5	4.360	271.8	85.99	20.70	50	0.25	650.23	0
650	Zearalenone (ppb)	0918	294.7	69.20	271.8	85.99	20.70	50	0.27	650.24	0
650	Zearalenone (ppb)	0590	296.0	14.00	271.8	85.99	20.70	50	0.28	650.06	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT	Your	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests	Z Score	Method	
650	Zearalenone (ppb)	0297	300.4	21.16	271.8	85.99	20.70	50	0.33	650.24	0
650	Zearalenone (ppb)	0563	300.5	15.00	271.8	85.99	20.70	50	0.33	650.20	0
650	Zearalenone (ppb)	2192	316.5	11.00	271.8	85.99	20.70	50	0.52	650.20	0
650	Zearalenone (ppb)	2327	326.5	5.000	271.8	85.99	20.70	50	0.64	650.24	0
650	Zearalenone (ppb)	2140	330.5	23.00	271.8	85.99	20.70	50	0.68	650.24	0
650	Zearalenone (ppb)	0010	336.5	3.000	271.8	85.99	20.70	50	0.75	650.24	0
650	Zearalenone (ppb)	2239	336.5	13.00	271.8	85.99	20.70	50	0.75	650.98	0
650	Zearalenone (ppb)	0529	337.2	2.400	271.8	85.99	20.70	50	0.76	650.01	0
650	Zearalenone (ppb)	0505	342.6	1.880	271.8	85.99	20.70	50	0.82	650.16	0
650	Zearalenone (ppb)	0015	347.4	28.22	271.8	85.99	20.70	50	0.88	650.16	0
650	Zearalenone (ppb)	0878	363.1	18.80	271.8	85.99	20.70	50	1.06	650.20	0
650	Zearalenone (ppb)	0035	367.3	13.00	271.8	85.99	20.70	50	1.11	650.07	0
650	Zearalenone (ppb)	0148	384.8	57.78	271.8	85.99	20.70	50	1.31	650.24	0
650	Zearalenone (ppb)	0596	420.9	28.40	271.8	85.99	20.70	50	1.73	650.98	0
650	Zearalenone (ppb)	2238	433.6	9.768	271.8	85.99	20.70	50	1.88	650.98	0
650	Zearalenone (ppb)	2268	435.0	10.00	271.8	85.99	20.70	50	1.90	650.99	0
650	Zearalenone (ppb)	0964	1,138	117.2	271.8	85.99	20.70	50	10.08	650.24	2
650	Zearalenone (ppb)	0572	< 500		271.8	85.99	20.70	50		650.24	5

- Note 1:** Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or between -2 and -3 (warning) and Green = Z <= 2 and >= -2 (OK at 95%). Horwitz SD's assigned based on Rob Mean only.
- Note 2:** Data Not Used Flags: 1=Rejected for duplicates too far apart, 2=Rejected as extreme outlier, 5=Reporting limit (<), 4=Zeros submitted as values and 3=Statistical problem. Flag 0 indicates data used in calculations.
- Note 3:** *Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). A zero range is not included in robust R-bar calculation.