



AAFCO
Association of American Feed Control Officials



Mycotoxin Contaminants Scheme

Dry Dog Feed

Test Material Code # 202562

Analyte Proficiency Testing Report

Labs Reporting: 67

Analytes Reported 15

Issue Date : 05/31/2025

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests			
600	Total Aflatoxin (ppb)	2196	5.504	0.0000	10.65	3.677	1.160	56	-1.40	600.20	0
600	Total Aflatoxin (ppb)	0202	6.450	2.600	10.65	3.677	1.160	56	-1.14	600.12	0
600	Total Aflatoxin (ppb)	2322	6.485	0.5300	10.65	3.677	1.160	56	-1.13	600.20	0
600	Total Aflatoxin (ppb)	0682	6.900	0.0000	10.65	3.677	1.160	56	-1.02	600.24	0
600	Total Aflatoxin (ppb)	0354	6.920	0.5400	10.65	3.677	1.160	56	-1.01	600.08	0
600	Total Aflatoxin (ppb)	0596	7.435	0.9300	10.65	3.677	1.160	56	-0.87	600.98	0
600	Total Aflatoxin (ppb)	2369	7.595	0.2700	10.65	3.677	1.160	56	-0.83	600.14	0
600	Total Aflatoxin (ppb)	2369	7.595	0.2700	10.65	3.677	1.160	56	-0.83	600.20	0
600	Total Aflatoxin (ppb)	0656	7.710	0.0000	10.65	3.677	1.160	56	-0.80	600.12	0
600	Total Aflatoxin (ppb)	2129	7.750	0.3000	10.65	3.677	1.160	56	-0.79	600.24	0
600	Total Aflatoxin (ppb)	0066	8.000	1.200	10.65	3.677	1.160	56	-0.72	600.13	0
600	Total Aflatoxin (ppb)	0010	8.300	2.400	10.65	3.677	1.160	56	-0.64	600.20	0
600	Total Aflatoxin (ppb)	0407	8.540	0.5000	10.65	3.677	1.160	56	-0.57	600.13	0
600	Total Aflatoxin (ppb)	2268	8.800	0.6000	10.65	3.677	1.160	56	-0.50	600.14	0
600	Total Aflatoxin (ppb)	2226	8.814	0.1250	10.65	3.677	1.160	56	-0.50	600.12	0
600	Total Aflatoxin (ppb)	2411	8.915	0.8900	10.65	3.677	1.160	56	-0.47	600.20	0
600	Total Aflatoxin (ppb)	0520	8.950	2.100	10.65	3.677	1.160	56	-0.46	600.14	0
600	Total Aflatoxin (ppb)	0354	9.156	0.9480	10.65	3.677	1.160	56	-0.41	600.14	0
600	Total Aflatoxin (ppb)	0354	9.156	0.9480	10.65	3.677	1.160	56	-0.41	600.22	0
600	Total Aflatoxin (ppb)	0034	9.175	1.350	10.65	3.677	1.160	56	-0.40	600.01	0
600	Total Aflatoxin (ppb)	0511	9.250	0.1000	10.65	3.677	1.160	56	-0.38	600.01	0
600	Total Aflatoxin (ppb)	0004	9.500	1.000	10.65	3.677	1.160	56	-0.31	600.14	0
600	Total Aflatoxin (ppb)	0563	9.660	1.100	10.65	3.677	1.160	56	-0.27	600.20	0
600	Total Aflatoxin (ppb)	2443	9.749	0.0000	10.65	3.677	1.160	56	-0.24	600.24	0
600	Total Aflatoxin (ppb)	0629	9.895	0.7300	10.65	3.677	1.160	56	-0.20	600.20	0
600	Total Aflatoxin (ppb)	2113	9.950	0.7000	10.65	3.677	1.160	56	-0.19	600.01	0
600	Total Aflatoxin (ppb)	0001	9.950	1.240	10.65	3.677	1.160	56	-0.19	600.20	0
600	Total Aflatoxin (ppb)	2096	10.10	7.800	10.65	3.677	1.160	56	-0.15	600.01	0
600	Total Aflatoxin (ppb)	0202	10.65	1.300	10.65	3.677	1.160	56	0.00	600.24	0
600	Total Aflatoxin (ppb)	2033	10.95	2.100	10.65	3.677	1.160	56	0.08	600.24	0
600	Total Aflatoxin (ppb)	0227	11.00	2.000	10.65	3.677	1.160	56	0.10	600.01	0
600	Total Aflatoxin (ppb)	2196	11.03	0.0000	10.65	3.677	1.160	56	0.10	600.08	0
600	Total Aflatoxin (ppb)	2192	11.10	0.8000	10.65	3.677	1.160	56	0.12	600.20	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests			
600	Total Aflatoxin (ppb)	0876	11.50	0.6000	10.65	3.677	1.160	56	0.23	600.01	0
600	Total Aflatoxin (ppb)	0027	11.53	0.3500	10.65	3.677	1.160	56	0.24	600.01	0
600	Total Aflatoxin (ppb)	0723	11.56	0.7400	10.65	3.677	1.160	56	0.25	600.20	0
600	Total Aflatoxin (ppb)	2022	11.60	1.400	10.65	3.677	1.160	56	0.26	600.01	0
600	Total Aflatoxin (ppb)	0014	11.85	1.500	10.65	3.677	1.160	56	0.33	600.01	0
600	Total Aflatoxin (ppb)	0051	12.15	3.700	10.65	3.677	1.160	56	0.41	600.20	0
600	Total Aflatoxin (ppb)	0297	12.30	1.600	10.65	3.677	1.160	56	0.45	600.24	0
600	Total Aflatoxin (ppb)	0208	12.35	1.900	10.65	3.677	1.160	56	0.46	600.01	0
600	Total Aflatoxin (ppb)	0918	12.50	5.950	10.65	3.677	1.160	56	0.50	600.01	0
600	Total Aflatoxin (ppb)	2022	12.70	0.3400	10.65	3.677	1.160	56	0.56	600.08	0
600	Total Aflatoxin (ppb)	2022	12.93	0.7200	10.65	3.677	1.160	56	0.62	600.11	0
600	Total Aflatoxin (ppb)	2147	13.00	2.000	10.65	3.677	1.160	56	0.64	600.04	0
600	Total Aflatoxin (ppb)	0098	13.01	0.9000	10.65	3.677	1.160	56	0.64	600.22	0
600	Total Aflatoxin (ppb)	2352	13.24	5.040	10.65	3.677	1.160	56	0.71	600.23	0
600	Total Aflatoxin (ppb)	0042	13.57	0.1400	10.65	3.677	1.160	56	0.79	600.20	0
600	Total Aflatoxin (ppb)	2239	14.06	1.600	10.65	3.677	1.160	56	0.93	600.98	0
600	Total Aflatoxin (ppb)	0353	14.50	1.000	10.65	3.677	1.160	56	1.05	600.24	0
600	Total Aflatoxin (ppb)	0674	15.20	0.0000	10.65	3.677	1.160	56	1.24	600.20	0
600	Total Aflatoxin (ppb)	0089	15.50	0.6000	10.65	3.677	1.160	56	1.32	600.02	0
600	Total Aflatoxin (ppb)	0148	15.81	1.700	10.65	3.677	1.160	56	1.40	600.24	0
600	Total Aflatoxin (ppb)	0278	16.05	0.5000	10.65	3.677	1.160	56	1.47	600.15	0
600	Total Aflatoxin (ppb)	2239	16.12	1.020	10.65	3.677	1.160	56	1.49	600.20	0
600	Total Aflatoxin (ppb)	2302	20.96	3.920	10.65	3.677	1.160	56	2.80	600.16	0
601	Aflatoxin B1 (ppb)	0047	1.300	0.1200	10.32	3.567	1.079	44	-2.53	601.99	0
601	Aflatoxin B1 (ppb)	0553	4.045	1.250	10.32	3.567	1.079	44	-1.76	601.23	0
601	Aflatoxin B1 (ppb)	2196	5.153	0.0000	10.32	3.567	1.079	44	-1.45	601.20	0
601	Aflatoxin B1 (ppb)	2322	5.430	0.7800	10.32	3.567	1.079	44	-1.37	601.20	0
601	Aflatoxin B1 (ppb)	0860	5.801	0.0000	10.32	3.567	1.079	44	-1.27	601.23	0
601	Aflatoxin B1 (ppb)	0242	6.000	2.000	10.32	3.567	1.079	44	-1.21	601.23	0
601	Aflatoxin B1 (ppb)	2369	6.830	0.2800	10.32	3.567	1.079	44	-0.98	601.20	0
601	Aflatoxin B1 (ppb)	2369	6.830	0.2800	10.32	3.567	1.079	44	-0.98	601.24	0
601	Aflatoxin B1 (ppb)	0682	6.900	0.0000	10.32	3.567	1.079	44	-0.96	601.23	0
601	Aflatoxin B1 (ppb)	0010	7.750	1.700	10.32	3.567	1.079	44	-0.72	601.20	0
601	Aflatoxin B1 (ppb)	2411	8.295	2.130	10.32	3.567	1.079	44	-0.57	601.20	0
601	Aflatoxin B1 (ppb)	0043	8.351	2.531	10.32	3.567	1.079	44	-0.55	601.23	0
601	Aflatoxin B1 (ppb)	0354	8.928	0.0120	10.32	3.567	1.079	44	-0.39	601.20	0
601	Aflatoxin B1 (ppb)	0354	8.928	0.0120	10.32	3.567	1.079	44	-0.39	601.24	0
601	Aflatoxin B1 (ppb)	2443	9.150	0.2740	10.32	3.567	1.079	44	-0.33	601.23	0
601	Aflatoxin B1 (ppb)	0918	9.250	2.500	10.32	3.567	1.079	44	-0.30	601.23	0
601	Aflatoxin B1 (ppb)	0629	9.450	0.4400	10.32	3.567	1.079	44	-0.24	601.20	0
601	Aflatoxin B1 (ppb)	0563	9.660	1.100	10.32	3.567	1.079	44	-0.18	601.20	0
601	Aflatoxin B1 (ppb)	0160	9.700	0.5000	10.32	3.567	1.079	44	-0.17	601.23	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT Z Score	Your	
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests		Method	Flag
601	Aflatoxin B1 (ppb)	2192	10.53	0.7000	10.32	3.567	1.079	44	0.06	601.20	0
601	Aflatoxin B1 (ppb)	2239	10.59	0.3800	10.32	3.567	1.079	44	0.08	601.98	0
601	Aflatoxin B1 (ppb)	0202	10.65	1.300	10.32	3.567	1.079	44	0.09	601.23	0
601	Aflatoxin B1 (ppb)	0009	10.74	0.0680	10.32	3.567	1.079	44	0.12	601.23	0
601	Aflatoxin B1 (ppb)	0723	10.77	0.6200	10.32	3.567	1.079	44	0.13	601.20	0
601	Aflatoxin B1 (ppb)	2238	10.89	1.220	10.32	3.567	1.079	44	0.16	601.98	0
601	Aflatoxin B1 (ppb)	2033	10.95	2.100	10.32	3.567	1.079	44	0.18	601.23	0
601	Aflatoxin B1 (ppb)	0675	11.00	0.2700	10.32	3.567	1.079	44	0.19	601.20	0
601	Aflatoxin B1 (ppb)	0051	11.15	3.500	10.32	3.567	1.079	44	0.23	601.20	0
601	Aflatoxin B1 (ppb)	2022	11.77	1.380	10.32	3.567	1.079	44	0.41	601.92	0
601	Aflatoxin B1 (ppb)	0870	11.99	0.6250	10.32	3.567	1.079	44	0.47	601.23	0
601	Aflatoxin B1 (ppb)	0098	12.01	0.9400	10.32	3.567	1.079	44	0.47	601.24	0
601	Aflatoxin B1 (ppb)	2239	12.25	0.0800	10.32	3.567	1.079	44	0.54	601.20	0
601	Aflatoxin B1 (ppb)	0297	12.28	1.610	10.32	3.567	1.079	44	0.55	601.23	0
601	Aflatoxin B1 (ppb)	2352	12.30	4.720	10.32	3.567	1.079	44	0.56	601.23	0
601	Aflatoxin B1 (ppb)	2022	12.64	0.9900	10.32	3.567	1.079	44	0.65	601.98	0
601	Aflatoxin B1 (ppb)	2022	13.15	1.050	10.32	3.567	1.079	44	0.79	601.91	0
601	Aflatoxin B1 (ppb)	0028	13.45	0.5500	10.32	3.567	1.079	44	0.88	601.23	0
601	Aflatoxin B1 (ppb)	0353	13.50	1.000	10.32	3.567	1.079	44	0.89	601.23	0
601	Aflatoxin B1 (ppb)	0913	13.50	1.200	10.32	3.567	1.079	44	0.89	601.23	0
601	Aflatoxin B1 (ppb)	0505	14.27	1.840	10.32	3.567	1.079	44	1.11	601.16	0
601	Aflatoxin B1 (ppb)	2327	14.95	0.3000	10.32	3.567	1.079	44	1.30	601.23	0
601	Aflatoxin B1 (ppb)	2302	15.04	0.0700	10.32	3.567	1.079	44	1.32	601.16	0
601	Aflatoxin B1 (ppb)	0689	15.10	3.600	10.32	3.567	1.079	44	1.34	601.16	0
601	Aflatoxin B1 (ppb)	0148	15.38	2.560	10.32	3.567	1.079	44	1.42	601.23	0
601	Aflatoxin B1 (ppb)	2140	< 10		10.32	3.567	1.079	44		601.23	5
602	Aflatoxin B2 (ppb)	2196	0.3510	0.0000	0.7553	0.2803	0.1410	20	-1.44	602.20	0
602	Aflatoxin B2 (ppb)	0629	0.4450	0.2900	0.7553	0.2803	0.1410	20	-1.11	602.20	0
602	Aflatoxin B2 (ppb)	0353	0.5150	0.0900	0.7553	0.2803	0.1410	20	-0.86	602.23	0
602	Aflatoxin B2 (ppb)	0010	0.5500	0.7000	0.7553	0.2803	0.1410	20	-0.73	602.20	0
602	Aflatoxin B2 (ppb)	2192	0.5700	0.1200	0.7553	0.2803	0.1410	20	-0.66	602.20	0
602	Aflatoxin B2 (ppb)	2443	0.5995	0.0490	0.7553	0.2803	0.1410	20	-0.56	602.23	0
602	Aflatoxin B2 (ppb)	0354	0.6840	0.0960	0.7553	0.2803	0.1410	20	-0.25	602.20	0
602	Aflatoxin B2 (ppb)	0354	0.6840	0.0960	0.7553	0.2803	0.1410	20	-0.25	602.24	0
602	Aflatoxin B2 (ppb)	2411	0.7400	0.2940	0.7553	0.2803	0.1410	20	-0.05	602.20	0
602	Aflatoxin B2 (ppb)	0870	0.7625	0.0250	0.7553	0.2803	0.1410	20	0.03	602.23	0
602	Aflatoxin B2 (ppb)	2369	0.7650	0.0100	0.7553	0.2803	0.1410	20	0.03	602.20	0
602	Aflatoxin B2 (ppb)	2369	0.7650	0.0100	0.7553	0.2803	0.1410	20	0.03	602.24	0
602	Aflatoxin B2 (ppb)	0723	0.7900	0.1200	0.7553	0.2803	0.1410	20	0.12	602.20	0
602	Aflatoxin B2 (ppb)	0051	0.8000	0.2000	0.7553	0.2803	0.1410	20	0.16	602.20	0
602	Aflatoxin B2 (ppb)	2327	0.8000	0.0000	0.7553	0.2803	0.1410	20	0.16	602.23	0
602	Aflatoxin B2 (ppb)	2352	0.9400	0.3200	0.7553	0.2803	0.1410	20	0.66	602.23	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests			
602	Aflatoxin B2 (ppb)	0098	1.000	0.0400	0.7553	0.2803	0.1410	20	0.87	602.24	0
602	Aflatoxin B2 (ppb)	0675	1.070	0.1600	0.7553	0.2803	0.1410	20	1.12	602.20	0
602	Aflatoxin B2 (ppb)	0964	1.215	0.1500	0.7553	0.2803	0.1410	20	1.64	602.23	0
602	Aflatoxin B2 (ppb)	0913	1.300	0.0000	0.7553	0.2803	0.1410	20	1.94	602.23	0
602	Aflatoxin B2 (ppb)	0028	0.0000	0.0000	0.7553	0.2803	0.1410	20		602.23	4
602	Aflatoxin B2 (ppb)	0297	0.0000	0.0000	0.7553	0.2803	0.1410	20		602.23	4
602	Aflatoxin B2 (ppb)	0918	0.6000	1.200	0.7553	0.2803	0.1410	20		602.23	4
602	Aflatoxin B2 (ppb)	0043	< 0.09		0.7553	0.2803	0.1410	20		602.23	5
602	Aflatoxin B2 (ppb)	0553	< 0.5		0.7553	0.2803	0.1410	20		602.23	5
602	Aflatoxin B2 (ppb)	0682	< 0.5		0.7553	0.2803	0.1410	20		602.23	5
602	Aflatoxin B2 (ppb)	0009	< 0.67		0.7553	0.2803	0.1410	20		602.23	5
602	Aflatoxin B2 (ppb)	0160	< 1		0.7553	0.2803	0.1410	20		602.23	5
602	Aflatoxin B2 (ppb)	0202	< 1		0.7553	0.2803	0.1410	20		602.23	5
602	Aflatoxin B2 (ppb)	0242	< 1		0.7553	0.2803	0.1410	20		602.23	5
602	Aflatoxin B2 (ppb)	2033	< 1		0.7553	0.2803	0.1410	20		602.23	5
602	Aflatoxin B2 (ppb)	0148	< 2.5		0.7553	0.2803	0.1410	20		602.23	5
602	Aflatoxin B2 (ppb)	0563	< 5		0.7553	0.2803	0.1410	20		602.20	5
602	Aflatoxin B2 (ppb)	2140	< 10		0.7553	0.2803	0.1410	20		602.23	5
603	Aflatoxin G1 (ppb)	0505	1.460	0.0000				3		603.16	0
603	Aflatoxin G1 (ppb)	2302	2.825	0.3500				3		603.16	0
603	Aflatoxin G1 (ppb)	0047	14.67	4.160				3		603.99	0
603	Aflatoxin G1 (ppb)	0675	0.0000	0.0000				3		603.20	4
603	Aflatoxin G1 (ppb)	0028	0.0000	0.0000				3		603.23	4
603	Aflatoxin G1 (ppb)	0297	0.0000	0.0000				3		603.23	4
603	Aflatoxin G1 (ppb)	0918	0.0000	0.0000				3		603.23	4
603	Aflatoxin G1 (ppb)	0629	< 0.25					3		603.20	5
603	Aflatoxin G1 (ppb)	0043	< 0.3					3		603.23	5
603	Aflatoxin G1 (ppb)	2411	< 0.33					3		603.20	5
603	Aflatoxin G1 (ppb)	0553	< 0.5					3		603.23	5
603	Aflatoxin G1 (ppb)	0682	< 0.5					3		603.23	5
603	Aflatoxin G1 (ppb)	2352	< 0.5					3		603.23	5
603	Aflatoxin G1 (ppb)	0051	< 0.6					3		603.20	5
603	Aflatoxin G1 (ppb)	0870	< 0.6					3		603.23	5
603	Aflatoxin G1 (ppb)	2327	< 0.6					3		603.23	5
603	Aflatoxin G1 (ppb)	0009	< 0.68					3		603.23	5
603	Aflatoxin G1 (ppb)	0160	< 1					3		603.23	5
603	Aflatoxin G1 (ppb)	0202	< 1					3		603.23	5
603	Aflatoxin G1 (ppb)	0242	< 1					3		603.23	5
603	Aflatoxin G1 (ppb)	0913	< 1					3		603.23	5
603	Aflatoxin G1 (ppb)	2033	< 1					3		603.23	5
603	Aflatoxin G1 (ppb)	0148	< 2.5					3		603.23	5
603	Aflatoxin G1 (ppb)	2140	< 10					3		603.23	5

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests			
603	Aflatoxin G1 (ppb)	0098	< 15					3		603.24	5
604	Aflatoxin G2 (ppb)	0964	0.3600	0.0200				1		604.23	0
604	Aflatoxin G2 (ppb)	0675	0.0000	0.0000				1		604.20	4
604	Aflatoxin G2 (ppb)	0028	0.0000	0.0000				1		604.23	4
604	Aflatoxin G2 (ppb)	0297	0.0000	0.0000				1		604.23	4
604	Aflatoxin G2 (ppb)	0918	0.0000	0.0000				1		604.23	4
604	Aflatoxin G2 (ppb)	0629	< 0.25					1		604.20	5
604	Aflatoxin G2 (ppb)	2411	< 0.33					1		604.20	5
604	Aflatoxin G2 (ppb)	0553	< 0.5					1		604.23	5
604	Aflatoxin G2 (ppb)	0682	< 0.5					1		604.23	5
604	Aflatoxin G2 (ppb)	0051	< 0.6					1		604.20	5
604	Aflatoxin G2 (ppb)	2327	< 0.6					1		604.23	5
604	Aflatoxin G2 (ppb)	0009	< 0.67					1		604.23	5
604	Aflatoxin G2 (ppb)	0160	< 1					1		604.23	5
604	Aflatoxin G2 (ppb)	0202	< 1					1		604.23	5
604	Aflatoxin G2 (ppb)	0242	< 1					1		604.23	5
604	Aflatoxin G2 (ppb)	0870	< 1					1		604.23	5
604	Aflatoxin G2 (ppb)	0913	< 1					1		604.23	5
604	Aflatoxin G2 (ppb)	2033	< 1					1		604.23	5
604	Aflatoxin G2 (ppb)	0148	< 3					1		604.23	5
604	Aflatoxin G2 (ppb)	0098	< 5					1		604.24	5
604	Aflatoxin G2 (ppb)	2140	< 10					1		604.23	5
610	Deoxynivalenol (ppb)	0596	0.5900	0.0200	514.7	160.0	40.48	61	-3.21	610.98	0
610	Deoxynivalenol (ppb)	0027	0.6350	0.1300	514.7	160.0	40.48	61	-3.21	610.01	0
610	Deoxynivalenol (ppb)	0553	184.5	93.00	514.7	160.0	40.48	61	-2.06	610.23	0
610	Deoxynivalenol (ppb)	0043	247.9	10.40	514.7	160.0	40.48	61	-1.67	610.23	0
610	Deoxynivalenol (ppb)	0354	295.7	65.27	514.7	160.0	40.48	61	-1.37	610.05	0
610	Deoxynivalenol (ppb)	2352	306.4	3.460	514.7	160.0	40.48	61	-1.30	610.23	0
610	Deoxynivalenol (ppb)	2147	335.0	110.0	514.7	160.0	40.48	61	-1.12	610.04	0
610	Deoxynivalenol (ppb)	0964	337.1	4.660	514.7	160.0	40.48	61	-1.11	610.23	0
610	Deoxynivalenol (ppb)	0047	352.7	62.85	514.7	160.0	40.48	61	-1.01	610.99	0
610	Deoxynivalenol (ppb)	2411	377.5	181.0	514.7	160.0	40.48	61	-0.86	610.20	0
610	Deoxynivalenol (ppb)	2129	390.0	16.00	514.7	160.0	40.48	61	-0.78	610.23	0
610	Deoxynivalenol (ppb)	0098	400.0	0.0000	514.7	160.0	40.48	61	-0.72	610.20	0
610	Deoxynivalenol (ppb)	2268	401.5	9.000	514.7	160.0	40.48	61	-0.71	610.22	0
610	Deoxynivalenol (ppb)	2033	425.0	30.00	514.7	160.0	40.48	61	-0.56	610.23	0
610	Deoxynivalenol (ppb)	2140	440.0	30.00	514.7	160.0	40.48	61	-0.47	610.23	0
610	Deoxynivalenol (ppb)	0870	440.6	61.50	514.7	160.0	40.48	61	-0.46	610.99	0
610	Deoxynivalenol (ppb)	2055	443.0	28.00	514.7	160.0	40.48	61	-0.45	610.23	0
610	Deoxynivalenol (ppb)	0918	449.4	32.66	514.7	160.0	40.48	61	-0.41	610.23	0
610	Deoxynivalenol (ppb)	0066	450.0	100.0	514.7	160.0	40.48	61	-0.40	610.08	0
610	Deoxynivalenol (ppb)	2196	453.8	0.0000	514.7	160.0	40.48	61	-0.38	610.20	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests			
610	Deoxynivalenol (ppb)	0689	458.0	106.0	514.7	160.0	40.48	61	-0.35	610.16	0
610	Deoxynivalenol (ppb)	0870	460.0	0.0000	514.7	160.0	40.48	61	-0.34	610.23	0
610	Deoxynivalenol (ppb)	0505	467.9	7.640	514.7	160.0	40.48	61	-0.29	610.16	0
610	Deoxynivalenol (ppb)	2369	468.5	13.00	514.7	160.0	40.48	61	-0.29	610.20	0
610	Deoxynivalenol (ppb)	0674	470.0	0.0000	514.7	160.0	40.48	61	-0.28	610.20	0
610	Deoxynivalenol (ppb)	0242	473.0	12.00	514.7	160.0	40.48	61	-0.26	610.23	0
610	Deoxynivalenol (ppb)	0353	476.5	11.00	514.7	160.0	40.48	61	-0.24	610.23	0
610	Deoxynivalenol (ppb)	0407	493.0	58.00	514.7	160.0	40.48	61	-0.14	610.07	0
610	Deoxynivalenol (ppb)	2327	498.2	4.900	514.7	160.0	40.48	61	-0.10	610.23	0
610	Deoxynivalenol (ppb)	0034	500.0	0.0000	514.7	160.0	40.48	61	-0.09	610.01	0
610	Deoxynivalenol (ppb)	0042	500.0	0.0000	514.7	160.0	40.48	61	-0.09	610.01	0
610	Deoxynivalenol (ppb)	2113	500.0	0.0000	514.7	160.0	40.48	61	-0.09	610.01	0
610	Deoxynivalenol (ppb)	2022	505.0	130.0	514.7	160.0	40.48	61	-0.06	610.01	0
610	Deoxynivalenol (ppb)	0682	510.0	0.0000	514.7	160.0	40.48	61	-0.03	610.23	0
610	Deoxynivalenol (ppb)	0051	511.5	3.000	514.7	160.0	40.48	61	-0.02	610.20	0
610	Deoxynivalenol (ppb)	2239	519.0	52.00	514.7	160.0	40.48	61	0.03	610.98	0
610	Deoxynivalenol (ppb)	0629	521.5	19.00	514.7	160.0	40.48	61	0.04	610.07	0
610	Deoxynivalenol (ppb)	2239	545.5	13.00	514.7	160.0	40.48	61	0.19	610.20	0
610	Deoxynivalenol (ppb)	0511	550.0	100.0	514.7	160.0	40.48	61	0.22	610.01	0
610	Deoxynivalenol (ppb)	0148	554.2	15.86	514.7	160.0	40.48	61	0.25	610.23	0
610	Deoxynivalenol (ppb)	0010	555.0	48.00	514.7	160.0	40.48	61	0.25	610.23	0
610	Deoxynivalenol (ppb)	0089	559.3	61.50	514.7	160.0	40.48	61	0.28	610.02	0
610	Deoxynivalenol (ppb)	2192	560.0	100.0	514.7	160.0	40.48	61	0.28	610.01	0
610	Deoxynivalenol (ppb)	2196	570.0	0.0000	514.7	160.0	40.48	61	0.35	610.05	0
610	Deoxynivalenol (ppb)	0028	588.2	118.8	514.7	160.0	40.48	61	0.46	610.23	0
610	Deoxynivalenol (ppb)	2238	618.0	4.070	514.7	160.0	40.48	61	0.65	610.98	0
610	Deoxynivalenol (ppb)	0918	623.5	17.00	514.7	160.0	40.48	61	0.68	610.01	0
610	Deoxynivalenol (ppb)	0001	624.5	17.00	514.7	160.0	40.48	61	0.69	610.20	0
610	Deoxynivalenol (ppb)	0009	644.7	55.88	514.7	160.0	40.48	61	0.81	610.23	0
610	Deoxynivalenol (ppb)	0208	655.0	90.00	514.7	160.0	40.48	61	0.88	610.01	0
610	Deoxynivalenol (ppb)	2022	690.0	20.00	514.7	160.0	40.48	61	1.10	610.07	0
610	Deoxynivalenol (ppb)	0160	700.0	0.0000	514.7	160.0	40.48	61	1.16	610.99	0
610	Deoxynivalenol (ppb)	0278	710.0	20.00	514.7	160.0	40.48	61	1.22	610.09	0
610	Deoxynivalenol (ppb)	0913	713.0	24.00	514.7	160.0	40.48	61	1.24	610.23	0
610	Deoxynivalenol (ppb)	2022	715.0	10.00	514.7	160.0	40.48	61	1.25	610.98	0
610	Deoxynivalenol (ppb)	2302	732.2	11.23	514.7	160.0	40.48	61	1.36	610.16	0
610	Deoxynivalenol (ppb)	0297	745.0	10.00	514.7	160.0	40.48	61	1.44	610.23	0
610	Deoxynivalenol (ppb)	2022	770.0	160.0	514.7	160.0	40.48	61	1.59	610.05	0
610	Deoxynivalenol (ppb)	0563	877.5	35.00	514.7	160.0	40.48	61	2.27	610.25	0
610	Deoxynivalenol (ppb)	0202	905.0	30.00	514.7	160.0	40.48	61	2.44	610.23	0
610	Deoxynivalenol (ppb)	0413	1,000	0.0000	514.7	160.0	40.48	61	3.03	610.20	0
610	Deoxynivalenol (ppb)	0656	< 222		514.7	160.0	40.48	61		610.07	5

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests			
610	Deoxynivalenol (ppb)	0014	< 500		514.7	160.0	40.48	61		610.01	5
610	Deoxynivalenol (ppb)	0227	< 500		514.7	160.0	40.48	61		610.01	5
610	Deoxynivalenol (ppb)	0202	< 500		514.7	160.0	40.48	61		610.06	5
620	Total Fumonisin (ppb)	0596	1.530	0.5400	1,112	338.5	95.91	42	-3.28	620.98	0
620	Total Fumonisin (ppb)	0202	1.720	0.3600	1,112	338.5	95.91	42	-3.28	620.09	0
620	Total Fumonisin (ppb)	0656	315.0	0.0000	1,112	338.5	95.91	42	-2.35	620.09	0
620	Total Fumonisin (ppb)	0511	650.0	100.0	1,112	338.5	95.91	42	-1.36	620.01	0
620	Total Fumonisin (ppb)	0689	671.5	99.00	1,112	338.5	95.91	42	-1.30	620.16	0
620	Total Fumonisin (ppb)	0089	730.0	20.00	1,112	338.5	95.91	42	-1.13	620.03	0
620	Total Fumonisin (ppb)	0042	750.0	100.0	1,112	338.5	95.91	42	-1.07	620.01	0
620	Total Fumonisin (ppb)	0918	754.5	117.0	1,112	338.5	95.91	42	-1.05	620.01	0
620	Total Fumonisin (ppb)	2411	768.0	8.000	1,112	338.5	95.91	42	-1.02	620.20	0
620	Total Fumonisin (ppb)	0098	800.0	400.0	1,112	338.5	95.91	42	-0.92	620.11	0
620	Total Fumonisin (ppb)	0682	810.0	0.0000	1,112	338.5	95.91	42	-0.89	620.23	0
620	Total Fumonisin (ppb)	0047	888.8	63.36	1,112	338.5	95.91	42	-0.66	620.99	0
620	Total Fumonisin (ppb)	0674	890.0	0.0000	1,112	338.5	95.91	42	-0.65	620.07	0
620	Total Fumonisin (ppb)	2369	908.0	2.540	1,112	338.5	95.91	42	-0.60	620.11	0
620	Total Fumonisin (ppb)	0520	945.0	50.00	1,112	338.5	95.91	42	-0.49	620.12	0
620	Total Fumonisin (ppb)	2192	960.0	0.0000	1,112	338.5	95.91	42	-0.45	620.01	0
620	Total Fumonisin (ppb)	0505	964.2	0.6600	1,112	338.5	95.91	42	-0.44	620.16	0
620	Total Fumonisin (ppb)	0034	997.5	85.00	1,112	338.5	95.91	42	-0.34	620.01	0
620	Total Fumonisin (ppb)	0227	1,000	0.0000	1,112	338.5	95.91	42	-0.33	620.07	0
620	Total Fumonisin (ppb)	0407	1,030	40.00	1,112	338.5	95.91	42	-0.24	620.09	0
620	Total Fumonisin (ppb)	2302	1,115	76.22	1,112	338.5	95.91	42	0.01	620.16	0
620	Total Fumonisin (ppb)	0010	1,128	36.00	1,112	338.5	95.91	42	0.05	620.23	0
620	Total Fumonisin (ppb)	2226	1,140	149.0	1,112	338.5	95.91	42	0.08	620.09	0
620	Total Fumonisin (ppb)	2268	1,150	100.0	1,112	338.5	95.91	42	0.11	620.11	0
620	Total Fumonisin (ppb)	2352	1,200	23.60	1,112	338.5	95.91	42	0.26	620.23	0
620	Total Fumonisin (ppb)	2129	1,227	178.0	1,112	338.5	95.91	42	0.34	620.23	0
620	Total Fumonisin (ppb)	0014	1,230	20.00	1,112	338.5	95.91	42	0.35	620.01	0
620	Total Fumonisin (ppb)	2443	1,250	0.0000	1,112	338.5	95.91	42	0.41	620.23	0
620	Total Fumonisin (ppb)	2022	1,280	160.0	1,112	338.5	95.91	42	0.50	620.08	0
620	Total Fumonisin (ppb)	0027	1,372	364.0	1,112	338.5	95.91	42	0.77	620.22	0
620	Total Fumonisin (ppb)	2033	1,400	0.0000	1,112	338.5	95.91	42	0.85	620.23	0
620	Total Fumonisin (ppb)	0297	1,425	170.0	1,112	338.5	95.91	42	0.93	620.23	0
620	Total Fumonisin (ppb)	0353	1,438	64.00	1,112	338.5	95.91	42	0.96	620.23	0
620	Total Fumonisin (ppb)	2022	1,450	140.0	1,112	338.5	95.91	42	1.00	620.07	0
620	Total Fumonisin (ppb)	0208	1,500	400.0	1,112	338.5	95.91	42	1.15	620.98	0
620	Total Fumonisin (ppb)	2239	1,515	14.00	1,112	338.5	95.91	42	1.19	620.20	0
620	Total Fumonisin (ppb)	0066	1,550	300.0	1,112	338.5	95.91	42	1.30	620.09	0
620	Total Fumonisin (ppb)	0202	1,590	20.00	1,112	338.5	95.91	42	1.41	620.23	0
620	Total Fumonisin (ppb)	2239	1,596	14.00	1,112	338.5	95.91	42	1.43	620.98	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests			
620	Total Fumonisin (ppb)	2022	1,665	90.00	1,112	338.5	95.91	42	1.64	620.01	0
620	Total Fumonisin (ppb)	2022	1,675	330.0	1,112	338.5	95.91	42	1.66	620.02	0
620	Total Fumonisin (ppb)	2147	1,750	500.0	1,112	338.5	95.91	42	1.89	620.04	0
620	Total Fumonisin (ppb)	2113	< 1000		1,112	338.5	95.91	42		620.01	5
621	Fumonsin B1 (ppb)	2411	587.0	20.00	838.4	257.3	34.67	27	-0.98	621.20	0
621	Fumonsin B1 (ppb)	0009	595.0	1.589	838.4	257.3	34.67	27	-0.95	621.23	0
621	Fumonsin B1 (ppb)	0913	609.5	21.00	838.4	257.3	34.67	27	-0.89	621.23	0
621	Fumonsin B1 (ppb)	0682	620.0	0.0000	838.4	257.3	34.67	27	-0.85	621.23	0
621	Fumonsin B1 (ppb)	0918	635.3	82.45	838.4	257.3	34.67	27	-0.79	621.23	0
621	Fumonsin B1 (ppb)	0043	656.1	0.3000	838.4	257.3	34.67	27	-0.71	621.23	0
621	Fumonsin B1 (ppb)	2369	711.1	3.130	838.4	257.3	34.67	27	-0.50	621.21	0
621	Fumonsin B1 (ppb)	0028	745.2	323.2	838.4	257.3	34.67	27	-0.36	621.23	0
621	Fumonsin B1 (ppb)	2140	747.5	53.00	838.4	257.3	34.67	27	-0.35	621.23	0
621	Fumonsin B1 (ppb)	0098	750.0	100.0	838.4	257.3	34.67	27	-0.34	621.21	0
621	Fumonsin B1 (ppb)	2238	751.4	18.52	838.4	257.3	34.67	27	-0.34	621.98	0
621	Fumonsin B1 (ppb)	2055	757.0	26.00	838.4	257.3	34.67	27	-0.32	621.23	0
621	Fumonsin B1 (ppb)	0051	758.5	15.00	838.4	257.3	34.67	27	-0.31	621.21	0
621	Fumonsin B1 (ppb)	0964	777.5	0.6700	838.4	257.3	34.67	27	-0.24	621.23	0
621	Fumonsin B1 (ppb)	2352	782.0	18.95	838.4	257.3	34.67	27	-0.22	621.23	0
621	Fumonsin B1 (ppb)	0010	801.0	6.000	838.4	257.3	34.67	27	-0.15	621.23	0
621	Fumonsin B1 (ppb)	2033	865.0	10.00	838.4	257.3	34.67	27	0.10	621.23	0
621	Fumonsin B1 (ppb)	0148	876.4	58.56	838.4	257.3	34.67	27	0.15	621.23	0
621	Fumonsin B1 (ppb)	0870	950.0	20.00	838.4	257.3	34.67	27	0.43	621.23	0
621	Fumonsin B1 (ppb)	0297	995.0	30.00	838.4	257.3	34.67	27	0.61	621.23	0
621	Fumonsin B1 (ppb)	0353	1,018	38.00	838.4	257.3	34.67	27	0.70	621.23	0
621	Fumonsin B1 (ppb)	2327	1,019	14.50	838.4	257.3	34.67	27	0.70	621.23	0
621	Fumonsin B1 (ppb)	2443	1,074	50.23	838.4	257.3	34.67	27	0.92	621.23	0
621	Fumonsin B1 (ppb)	0202	1,100	0.0000	838.4	257.3	34.67	27	1.02	621.23	0
621	Fumonsin B1 (ppb)	0160	1,150	300.0	838.4	257.3	34.67	27	1.21	621.23	0
621	Fumonsin B1 (ppb)	0553	1,399	208.0	838.4	257.3	34.67	27	2.18	621.23	0
621	Fumonsin B1 (ppb)	0563	1,466	34.78	838.4	257.3	34.67	27	2.44	621.20	0
622	Fumonisin B2 (ppb)	0160	150.0	100.0	257.7	81.63	17.32	25	-1.32	622.23	0
622	Fumonisin B2 (ppb)	0043	166.6	9.400	257.7	81.63	17.32	25	-1.12	622.23	0
622	Fumonisin B2 (ppb)	2443	175.5	0.0000	257.7	81.63	17.32	25	-1.01	622.23	0
622	Fumonisin B2 (ppb)	2411	181.0	12.00	257.7	81.63	17.32	25	-0.94	622.20	0
622	Fumonisin B2 (ppb)	0682	190.0	0.0000	257.7	81.63	17.32	25	-0.83	622.23	0
622	Fumonisin B2 (ppb)	2369	197.0	0.5900	257.7	81.63	17.32	25	-0.74	622.21	0
622	Fumonisin B2 (ppb)	2140	198.0	42.00	257.7	81.63	17.32	25	-0.73	622.23	0
622	Fumonisin B2 (ppb)	0098	200.0	0.0000	257.7	81.63	17.32	25	-0.71	622.21	0
622	Fumonisin B2 (ppb)	0010	206.5	3.000	257.7	81.63	17.32	25	-0.63	622.23	0
622	Fumonisin B2 (ppb)	0913	222.5	1.000	257.7	81.63	17.32	25	-0.43	622.23	0
622	Fumonisin B2 (ppb)	0009	238.9	5.978	257.7	81.63	17.32	25	-0.23	622.23	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT Z Score	Your	
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests		Method	Flag
622	Fumonisin B2 (ppb)	0051	251.0	0.0000	257.7	81.63	17.32	25	-0.08	622.21	0
622	Fumonisin B2 (ppb)	2055	259.0	34.00	257.7	81.63	17.32	25	0.02	622.23	0
622	Fumonisin B2 (ppb)	0553	269.4	89.90	257.7	81.63	17.32	25	0.14	622.23	0
622	Fumonisin B2 (ppb)	0028	271.0	8.111	257.7	81.63	17.32	25	0.16	622.23	0
622	Fumonisin B2 (ppb)	0297	275.0	30.00	257.7	81.63	17.32	25	0.21	622.23	0
622	Fumonisin B2 (ppb)	2352	293.0	5.750	257.7	81.63	17.32	25	0.43	622.23	0
622	Fumonisin B2 (ppb)	2327	296.8	0.7000	257.7	81.63	17.32	25	0.48	622.23	0
622	Fumonisin B2 (ppb)	0148	308.1	8.870	257.7	81.63	17.32	25	0.62	622.23	0
622	Fumonisin B2 (ppb)	0353	321.0	2.000	257.7	81.63	17.32	25	0.78	622.23	0
622	Fumonisin B2 (ppb)	0918	325.6	5.560	257.7	81.63	17.32	25	0.83	622.23	0
622	Fumonisin B2 (ppb)	0870	340.0	40.00	257.7	81.63	17.32	25	1.01	622.23	0
622	Fumonisin B2 (ppb)	0202	360.0	40.00	257.7	81.63	17.32	25	1.25	622.23	0
622	Fumonisin B2 (ppb)	0964	380.9	2.040	257.7	81.63	17.32	25	1.51	622.23	0
622	Fumonisin B2 (ppb)	2033	400.0	0.0000	257.7	81.63	17.32	25	1.74	622.23	0
623	Fumonisin B3 (ppb)	0160	90.50	19.00	127.7	41.24	14.37	12	-0.90	623.23	0
623	Fumonisin B3 (ppb)	0353	107.0	2.000	127.7	41.24	14.37	12	-0.50	623.23	0
623	Fumonisin B3 (ppb)	0009	110.5	1.028	127.7	41.24	14.37	12	-0.42	623.23	0
623	Fumonisin B3 (ppb)	0870	120.0	0.0000	127.7	41.24	14.37	12	-0.19	623.23	0
623	Fumonisin B3 (ppb)	0010	120.5	45.00	127.7	41.24	14.37	12	-0.18	623.23	0
623	Fumonisin B3 (ppb)	2352	124.8	10.41	127.7	41.24	14.37	12	-0.07	623.23	0
623	Fumonisin B3 (ppb)	0918	129.8	15.96	127.7	41.24	14.37	12	0.05	623.23	0
623	Fumonisin B3 (ppb)	0202	130.0	20.00	127.7	41.24	14.37	12	0.06	623.23	0
623	Fumonisin B3 (ppb)	2033	135.0	10.00	127.7	41.24	14.37	12	0.18	623.23	0
623	Fumonisin B3 (ppb)	0553	144.7	7.300	127.7	41.24	14.37	12	0.41	623.23	0
623	Fumonisin B3 (ppb)	0297	155.0	170.0	127.7	41.24	14.37	12	0.66	623.23	0
623	Fumonisin B3 (ppb)	0964	261.1	2.990	127.7	41.24	14.37	12	3.23	623.23	0
623	Fumonisin B3 (ppb)	0913	< 100		127.7	41.24	14.37	12		623.23	5
623	Fumonisin B3 (ppb)	2327	< 160		127.7	41.24	14.37	12		623.23	5
630	Ochratoxin A (ppb)	0656	2.940	0.0000	13.61	4.670	1.585	51	-2.29	630.05	0
630	Ochratoxin A (ppb)	0227	3.000	0.0000	13.61	4.670	1.585	51	-2.27	630.01	0
630	Ochratoxin A (ppb)	0918	3.470	0.4000	13.61	4.670	1.585	51	-2.17	630.01	0
630	Ochratoxin A (ppb)	0553	5.420	0.3000	13.61	4.670	1.585	51	-1.75	630.23	0
630	Ochratoxin A (ppb)	0918	6.700	0.6000	13.61	4.670	1.585	51	-1.48	630.23	0
630	Ochratoxin A (ppb)	0689	7.450	1.700	13.61	4.670	1.585	51	-1.32	630.16	0
630	Ochratoxin A (ppb)	0596	8.950	3.900	13.61	4.670	1.585	51	-1.00	630.98	0
630	Ochratoxin A (ppb)	2226	8.994	0.4440	13.61	4.670	1.585	51	-0.99	630.05	0
630	Ochratoxin A (ppb)	0629	9.330	0.1200	13.61	4.670	1.585	51	-0.92	630.20	0
630	Ochratoxin A (ppb)	2033	10.20	1.600	13.61	4.670	1.585	51	-0.73	630.23	0
630	Ochratoxin A (ppb)	0870	10.24	1.120	13.61	4.670	1.585	51	-0.72	630.23	0
630	Ochratoxin A (ppb)	0723	10.97	1.690	13.61	4.670	1.585	51	-0.57	630.20	0
630	Ochratoxin A (ppb)	2140	11.00	4.000	13.61	4.670	1.585	51	-0.56	630.23	0
630	Ochratoxin A (ppb)	2022	11.93	0.7600	13.61	4.670	1.585	51	-0.36	630.03	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT Z Score	Your	
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests		Method	Flag
630	Ochratoxin A (ppb)	0297	12.00	4.000	13.61	4.670	1.585	51	-0.35	630.23	0
630	Ochratoxin A (ppb)	0505	12.30	0.9200	13.61	4.670	1.585	51	-0.28	630.16	0
630	Ochratoxin A (ppb)	2352	12.37	0.5200	13.61	4.670	1.585	51	-0.27	630.23	0
630	Ochratoxin A (ppb)	2238	12.65	1.700	13.61	4.670	1.585	51	-0.21	630.98	0
630	Ochratoxin A (ppb)	2302	12.74	4.520	13.61	4.670	1.585	51	-0.19	630.16	0
630	Ochratoxin A (ppb)	0964	12.79	1.100	13.61	4.670	1.585	51	-0.18	630.23	0
630	Ochratoxin A (ppb)	0047	12.80	4.740	13.61	4.670	1.585	51	-0.17	630.99	0
630	Ochratoxin A (ppb)	2055	12.85	1.100	13.61	4.670	1.585	51	-0.16	630.23	0
630	Ochratoxin A (ppb)	0089	12.95	0.3000	13.61	4.670	1.585	51	-0.14	630.01	0
630	Ochratoxin A (ppb)	0563	13.23	2.170	13.61	4.670	1.585	51	-0.08	630.20	0
630	Ochratoxin A (ppb)	0913	13.35	1.500	13.61	4.670	1.585	51	-0.06	630.23	0
630	Ochratoxin A (ppb)	0354	13.47	4.040	13.61	4.670	1.585	51	-0.03	630.03	0
630	Ochratoxin A (ppb)	0160	13.50	3.000	13.61	4.670	1.585	51	-0.02	630.23	0
630	Ochratoxin A (ppb)	0675	14.01	0.0500	13.61	4.670	1.585	51	0.08	630.20	0
630	Ochratoxin A (ppb)	0354	14.08	1.266	13.61	4.670	1.585	51	0.10	630.06	0
630	Ochratoxin A (ppb)	0354	14.08	1.266	13.61	4.670	1.585	51	0.10	630.20	0
630	Ochratoxin A (ppb)	0148	14.49	0.7000	13.61	4.670	1.585	51	0.19	630.23	0
630	Ochratoxin A (ppb)	0353	14.50	5.000	13.61	4.670	1.585	51	0.19	630.23	0
630	Ochratoxin A (ppb)	0010	14.70	4.400	13.61	4.670	1.585	51	0.23	630.23	0
630	Ochratoxin A (ppb)	0520	15.00	2.000	13.61	4.670	1.585	51	0.30	630.06	0
630	Ochratoxin A (ppb)	0682	15.40	0.0000	13.61	4.670	1.585	51	0.38	630.23	0
630	Ochratoxin A (ppb)	2239	15.42	0.8700	13.61	4.670	1.585	51	0.39	630.98	0
630	Ochratoxin A (ppb)	0043	15.76	0.1900	13.61	4.670	1.585	51	0.46	630.23	0
630	Ochratoxin A (ppb)	2022	16.04	0.0900	13.61	4.670	1.585	51	0.52	630.01	0
630	Ochratoxin A (ppb)	2022	16.13	0.5900	13.61	4.670	1.585	51	0.54	630.04	0
630	Ochratoxin A (ppb)	0674	16.79	0.0000	13.61	4.670	1.585	51	0.68	630.20	0
630	Ochratoxin A (ppb)	0098	17.05	2.700	13.61	4.670	1.585	51	0.74	630.20	0
630	Ochratoxin A (ppb)	0407	17.42	2.660	13.61	4.670	1.585	51	0.82	630.04	0
630	Ochratoxin A (ppb)	0860	17.63	0.0000	13.61	4.670	1.585	51	0.86	630.23	0
630	Ochratoxin A (ppb)	2129	18.15	1.100	13.61	4.670	1.585	51	0.97	630.23	0
630	Ochratoxin A (ppb)	0051	18.30	0.4000	13.61	4.670	1.585	51	1.00	630.20	0
630	Ochratoxin A (ppb)	2369	18.45	0.8000	13.61	4.670	1.585	51	1.04	630.06	0
630	Ochratoxin A (ppb)	2369	18.45	0.8000	13.61	4.670	1.585	51	1.04	630.20	0
630	Ochratoxin A (ppb)	0202	20.25	0.7000	13.61	4.670	1.585	51	1.42	630.23	0
630	Ochratoxin A (ppb)	0066	20.75	3.300	13.61	4.670	1.585	51	1.53	630.05	0
630	Ochratoxin A (ppb)	2327	21.95	1.900	13.61	4.670	1.585	51	1.79	630.23	0
630	Ochratoxin A (ppb)	2443	32.18	1.064	13.61	4.670	1.585	51	3.98	630.23	0
630	Ochratoxin A (ppb)	2322	58.78	8.600	13.61	4.670	1.585	51	9.67	630.20	2
640	T-2 Toxin (ppb)	0553	31.65	8.100	105.1	34.13	10.08	25	-2.15	640.23	0
640	T-2 Toxin (ppb)	2238	51.02	3.540	105.1	34.13	10.08	25	-1.59	640.98	0
640	T-2 Toxin (ppb)	0674	57.68	0.0000	105.1	34.13	10.08	25	-1.39	640.03	0
640	T-2 Toxin (ppb)	0918	58.47	8.940	105.1	34.13	10.08	25	-1.37	640.23	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests			
640	T-2 Toxin (ppb)	0682	72.10	0.0000	105.1	34.13	10.08	25	-0.97	640.23	0
640	T-2 Toxin (ppb)	2352	74.52	7.610	105.1	34.13	10.08	25	-0.90	640.23	0
640	T-2 Toxin (ppb)	2129	85.65	1.300	105.1	34.13	10.08	25	-0.57	640.23	0
640	T-2 Toxin (ppb)	0202	87.30	8.600	105.1	34.13	10.08	25	-0.52	640.23	0
640	T-2 Toxin (ppb)	2327	88.25	3.500	105.1	34.13	10.08	25	-0.50	640.23	0
640	T-2 Toxin (ppb)	2055	95.70	16.60	105.1	34.13	10.08	25	-0.28	640.23	0
640	T-2 Toxin (ppb)	0160	105.0	10.00	105.1	34.13	10.08	25	0.00	640.23	0
640	T-2 Toxin (ppb)	0051	109.5	1.000	105.1	34.13	10.08	25	0.13	640.23	0
640	T-2 Toxin (ppb)	2140	109.5	31.00	105.1	34.13	10.08	25	0.13	640.23	0
640	T-2 Toxin (ppb)	0148	110.8	0.9900	105.1	34.13	10.08	25	0.17	640.23	0
640	T-2 Toxin (ppb)	0964	110.9	13.01	105.1	34.13	10.08	25	0.17	640.23	0
640	T-2 Toxin (ppb)	0913	115.7	36.40	105.1	34.13	10.08	25	0.31	640.23	0
640	T-2 Toxin (ppb)	0870	125.3	6.380	105.1	34.13	10.08	25	0.59	640.23	0
640	T-2 Toxin (ppb)	0353	130.5	3.000	105.1	34.13	10.08	25	0.74	640.23	0
640	T-2 Toxin (ppb)	0297	133.5	13.00	105.1	34.13	10.08	25	0.83	640.23	0
640	T-2 Toxin (ppb)	2369	134.1	1.130	105.1	34.13	10.08	25	0.85	640.23	0
640	T-2 Toxin (ppb)	2033	135.0	30.00	105.1	34.13	10.08	25	0.87	640.23	0
640	T-2 Toxin (ppb)	0407	138.8	19.18	105.1	34.13	10.08	25	0.99	640.05	0
640	T-2 Toxin (ppb)	2239	139.3	3.900	105.1	34.13	10.08	25	1.00	640.98	0
640	T-2 Toxin (ppb)	0563	151.5	7.000	105.1	34.13	10.08	25	1.36	640.25	0
640	T-2 Toxin (ppb)	0066	157.8	27.40	105.1	34.13	10.08	25	1.54	640.03	0
640	T-2 Toxin (ppb)	0656	< 20		105.1	34.13	10.08	25		640.05	5
641	HT-2 Toxin (ppb)	0297	7.550	4.640	17.87	6.085	3.202	9	-1.70	641.23	0
641	HT-2 Toxin (ppb)	0353	10.00	2.000	17.87	6.085	3.202	9	-1.29	641.23	0
641	HT-2 Toxin (ppb)	0202	14.50	0.6000	17.87	6.085	3.202	9	-0.55	641.23	0
641	HT-2 Toxin (ppb)	2369	15.53	0.3200	17.87	6.085	3.202	9	-0.38	641.23	0
641	HT-2 Toxin (ppb)	0051	20.00	6.000	17.87	6.085	3.202	9	0.35	641.23	0
641	HT-2 Toxin (ppb)	0148	20.73	2.720	17.87	6.085	3.202	9	0.47	641.23	0
641	HT-2 Toxin (ppb)	0964	21.60	0.4400	17.87	6.085	3.202	9	0.61	641.23	0
641	HT-2 Toxin (ppb)	0913	25.40	4.400	17.87	6.085	3.202	9	1.24	641.23	0
641	HT-2 Toxin (ppb)	2033	25.50	13.00	17.87	6.085	3.202	9	1.25	641.23	0
641	HT-2 Toxin (ppb)	0918	0.0000	0.0000	17.87	6.085	3.202	9		641.23	4
641	HT-2 Toxin (ppb)	2352	< 3		17.87	6.085	3.202	9		641.23	5
641	HT-2 Toxin (ppb)	2327	< 15		17.87	6.085	3.202	9		641.23	5
641	HT-2 Toxin (ppb)	0870	< 25		17.87	6.085	3.202	9		641.23	5
641	HT-2 Toxin (ppb)	2140	< 200		17.87	6.085	3.202	9		641.23	5
642	T-2 Toxin + HT-2 Toxin (ppb)	0596	61.75	21.50	125.3	40.48	10.47	20	-1.57	642.98	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0629	91.30	9.400	125.3	40.48	10.47	20	-0.84	642.98	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0689	93.50	9.000	125.3	40.48	10.47	20	-0.79	642.16	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0505	94.54	7.320	125.3	40.48	10.47	20	-0.76	642.16	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0407	97.81	18.16	125.3	40.48	10.47	20	-0.68	642.99	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0202	101.9	8.200	125.3	40.48	10.47	20	-0.58	642.04	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT Z Score	Your	
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests		Method	Flag
642	T-2 Toxin + HT-2 Toxin (ppb)	0202	112.0	12.00	125.3	40.48	10.47	20	-0.33	642.98	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0227	114.0	6.000	125.3	40.48	10.47	20	-0.28	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2192	118.0	10.00	125.3	40.48	10.47	20	-0.18	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0353	127.0	4.000	125.3	40.48	10.47	20	0.04	642.04	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0051	129.5	5.000	125.3	40.48	10.47	20	0.10	642.04	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0047	129.8	36.54	125.3	40.48	10.47	20	0.11	642.99	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0297	138.4	3.230	125.3	40.48	10.47	20	0.32	642.04	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0089	145.9	3.000	125.3	40.48	10.47	20	0.51	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2302	147.3	37.47	125.3	40.48	10.47	20	0.54	642.16	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2369	149.6	0.8100	125.3	40.48	10.47	20	0.60	642.04	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2022	151.9	13.78	125.3	40.48	10.47	20	0.66	642.99	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2022	155.1	5.440	125.3	40.48	10.47	20	0.74	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2022	157.7	2.170	125.3	40.48	10.47	20	0.80	642.98	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0918	232.8	150.6	125.3	40.48	10.47	20	2.66	642.21	0
650	Zearalenone (ppb)	0674	23.94	0.0000	89.76	29.26	9.381	54	-2.25	650.20	0
650	Zearalenone (ppb)	0596	26.00	5.200	89.76	29.26	9.381	54	-2.18	650.98	0
650	Zearalenone (ppb)	0553	41.65	3.700	89.76	29.26	9.381	54	-1.64	650.24	0
650	Zearalenone (ppb)	2129	44.70	0.8000	89.76	29.26	9.381	54	-1.54	650.24	0
650	Zearalenone (ppb)	0043	56.47	5.870	89.76	29.26	9.381	54	-1.14	650.24	0
650	Zearalenone (ppb)	2147	57.00	6.000	89.76	29.26	9.381	54	-1.12	650.02	0
650	Zearalenone (ppb)	0160	57.50	13.00	89.76	29.26	9.381	54	-1.10	650.24	0
650	Zearalenone (ppb)	0242	61.00	10.00	89.76	29.26	9.381	54	-0.98	650.24	0
650	Zearalenone (ppb)	0918	62.40	2.600	89.76	29.26	9.381	54	-0.93	650.01	0
650	Zearalenone (ppb)	0918	67.14	11.37	89.76	29.26	9.381	54	-0.77	650.24	0
650	Zearalenone (ppb)	0354	67.48	2.870	89.76	29.26	9.381	54	-0.76	650.04	0
650	Zearalenone (ppb)	0089	67.55	18.50	89.76	29.26	9.381	54	-0.76	650.01	0
650	Zearalenone (ppb)	0505	68.62	0.6800	89.76	29.26	9.381	54	-0.72	650.16	0
650	Zearalenone (ppb)	2369	68.99	1.280	89.76	29.26	9.381	54	-0.71	650.20	0
650	Zearalenone (ppb)	2369	68.99	1.280	89.76	29.26	9.381	54	-0.71	650.21	0
650	Zearalenone (ppb)	2140	74.50	25.00	89.76	29.26	9.381	54	-0.52	650.24	0
650	Zearalenone (ppb)	2352	74.82	0.3700	89.76	29.26	9.381	54	-0.51	650.24	0
650	Zearalenone (ppb)	2033	76.50	15.00	89.76	29.26	9.381	54	-0.45	650.24	0
650	Zearalenone (ppb)	0870	76.88	5.181	89.76	29.26	9.381	54	-0.44	650.01	0
650	Zearalenone (ppb)	0656	80.00	0.0000	89.76	29.26	9.381	54	-0.33	650.06	0
650	Zearalenone (ppb)	0563	80.40	4.980	89.76	29.26	9.381	54	-0.32	650.20	0
650	Zearalenone (ppb)	0689	81.00	14.00	89.76	29.26	9.381	54	-0.30	650.16	0
650	Zearalenone (ppb)	2226	83.42	5.372	89.76	29.26	9.381	54	-0.22	650.06	0
650	Zearalenone (ppb)	0066	84.10	26.80	89.76	29.26	9.381	54	-0.19	650.07	0
650	Zearalenone (ppb)	0354	84.16	2.364	89.76	29.26	9.381	54	-0.19	650.20	0
650	Zearalenone (ppb)	0354	84.16	2.364	89.76	29.26	9.381	54	-0.19	650.21	0
650	Zearalenone (ppb)	0047	89.22	47.24	89.76	29.26	9.381	54	-0.02	650.99	0
650	Zearalenone (ppb)	0913	91.45	1.900	89.76	29.26	9.381	54	0.06	650.24	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests			
650	Zearalenone (ppb)	0629	92.81	3.790	89.76	29.26	9.381	54	0.10	650.20	0
650	Zearalenone (ppb)	2196	95.46	0.0000	89.76	29.26	9.381	54	0.19	650.04	0
650	Zearalenone (ppb)	2196	97.29	0.0000	89.76	29.26	9.381	54	0.26	650.20	0
650	Zearalenone (ppb)	2268	97.50	5.000	89.76	29.26	9.381	54	0.26	650.99	0
650	Zearalenone (ppb)	0353	98.00	0.0000	89.76	29.26	9.381	54	0.28	650.23	0
650	Zearalenone (ppb)	0413	100.0	0.0000	89.76	29.26	9.381	54	0.35	650.20	0
650	Zearalenone (ppb)	0027	101.2	19.26	89.76	29.26	9.381	54	0.39	650.01	0
650	Zearalenone (ppb)	0051	102.0	18.00	89.76	29.26	9.381	54	0.42	650.20	0
650	Zearalenone (ppb)	2022	104.2	10.25	89.76	29.26	9.381	54	0.49	650.04	0
650	Zearalenone (ppb)	0098	105.0	10.00	89.76	29.26	9.381	54	0.52	650.20	0
650	Zearalenone (ppb)	2192	105.5	15.00	89.76	29.26	9.381	54	0.54	650.20	0
650	Zearalenone (ppb)	2239	107.0	8.000	89.76	29.26	9.381	54	0.59	650.20	0
650	Zearalenone (ppb)	2239	110.0	8.000	89.76	29.26	9.381	54	0.69	650.98	0
650	Zearalenone (ppb)	0202	111.0	16.00	89.76	29.26	9.381	54	0.73	650.24	0
650	Zearalenone (ppb)	0148	114.8	1.820	89.76	29.26	9.381	54	0.86	650.24	0
650	Zearalenone (ppb)	0407	114.8	20.93	89.76	29.26	9.381	54	0.86	650.07	0
650	Zearalenone (ppb)	0520	115.0	30.00	89.76	29.26	9.381	54	0.86	650.99	0
650	Zearalenone (ppb)	2022	115.4	8.330	89.76	29.26	9.381	54	0.88	650.01	0
650	Zearalenone (ppb)	0010	115.5	5.000	89.76	29.26	9.381	54	0.88	650.24	0
650	Zearalenone (ppb)	2238	117.6	5.790	89.76	29.26	9.381	54	0.95	650.98	0
650	Zearalenone (ppb)	2443	125.5	0.0000	89.76	29.26	9.381	54	1.22	650.24	0
650	Zearalenone (ppb)	2327	127.0	16.40	89.76	29.26	9.381	54	1.27	650.24	0
650	Zearalenone (ppb)	2302	127.2	29.56	89.76	29.26	9.381	54	1.28	650.16	0
650	Zearalenone (ppb)	0297	127.5	3.000	89.76	29.26	9.381	54	1.29	650.24	0
650	Zearalenone (ppb)	0682	129.8	0.0000	89.76	29.26	9.381	54	1.37	650.24	0
650	Zearalenone (ppb)	0964	150.7	15.10	89.76	29.26	9.381	54	2.08	650.24	0
650	Zearalenone (ppb)	0202	< 100		89.76	29.26	9.381	54		650.06	5

Note 1: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or between -2 and -3 (warning) and Green = Z <= 2 and >= -2 (OK at 95%). Horwitz SD's assigned based on Rob Mean only.

Note 2: Data Not Used Flags: 1=Rejected for duplicates too far apart, 2=Rejected as extreme outlier, 5=Reporting limit (<), 4=Zeros submitted as values and 3=Statistical problem. Flag 0 indicates data used in calculations.

Note 3: *Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). A zero range is not included in robust R-bar calculation.