



AAFCO
Association of American Feed Control Officials



Mycotoxin Contaminants Scheme

Livestock Feed

Test Material Code # 202564

Analyte Proficiency Testing Report

Labs Reporting: 65

Analytes Reported 15

Issue Date : 11/30/2025

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests			
600	Total Aflatoxin (ppb)	0066	4.400	0.4000	11.75	4.046	1.192	59	-1.82	600.13	0
600	Total Aflatoxin (ppb)	2326	4.500	1.000	11.75	4.046	1.192	59	-1.79	600.23	0
600	Total Aflatoxin (ppb)	2066	5.575	0.3100	11.75	4.046	1.192	59	-1.53	600.12	0
600	Total Aflatoxin (ppb)	2192	6.150	0.1000	11.75	4.046	1.192	59	-1.38	600.20	0
600	Total Aflatoxin (ppb)	0278	6.700	0.6000	11.75	4.046	1.192	59	-1.25	600.15	0
600	Total Aflatoxin (ppb)	2369	7.385	0.2900	11.75	4.046	1.192	59	-1.08	600.14	0
600	Total Aflatoxin (ppb)	2369	7.385	0.2900	11.75	4.046	1.192	59	-1.08	600.20	0
600	Total Aflatoxin (ppb)	0674	7.480	0.0000	11.75	4.046	1.192	59	-1.05	600.20	0
600	Total Aflatoxin (ppb)	2066	7.765	0.0500	11.75	4.046	1.192	59	-0.98	600.20	0
600	Total Aflatoxin (ppb)	0004	8.300	0.2000	11.75	4.046	1.192	59	-0.85	600.14	0
600	Total Aflatoxin (ppb)	2196	8.320	0.0000	11.75	4.046	1.192	59	-0.85	600.20	0
600	Total Aflatoxin (ppb)	0876	8.550	4.500	11.75	4.046	1.192	59	-0.79	600.01	0
600	Total Aflatoxin (ppb)	0682	9.100	0.0000	11.75	4.046	1.192	59	-0.65	600.24	0
600	Total Aflatoxin (ppb)	2268	9.360	0.6800	11.75	4.046	1.192	59	-0.59	600.14	0
600	Total Aflatoxin (ppb)	0202	9.440	0.4800	11.75	4.046	1.192	59	-0.57	600.24	0
600	Total Aflatoxin (ppb)	0505	9.650	0.1000	11.75	4.046	1.192	59	-0.52	600.22	0
600	Total Aflatoxin (ppb)	2033	9.750	2.500	11.75	4.046	1.192	59	-0.49	600.24	0
600	Total Aflatoxin (ppb)	0590	10.05	0.1000	11.75	4.046	1.192	59	-0.42	600.01	0
600	Total Aflatoxin (ppb)	0027	10.52	0.1300	11.75	4.046	1.192	59	-0.30	600.01	0
600	Total Aflatoxin (ppb)	0010	10.65	0.9000	11.75	4.046	1.192	59	-0.27	600.20	0
600	Total Aflatoxin (ppb)	2199	10.80	2.400	11.75	4.046	1.192	59	-0.23	600.14	0
600	Total Aflatoxin (ppb)	2022	10.91	0.0400	11.75	4.046	1.192	59	-0.21	600.11	0
600	Total Aflatoxin (ppb)	0554	10.95	0.5000	11.75	4.046	1.192	59	-0.20	600.01	0
600	Total Aflatoxin (ppb)	0033	11.00	1.600	11.75	4.046	1.192	59	-0.18	600.98	0
600	Total Aflatoxin (ppb)	0511	11.10	0.8000	11.75	4.046	1.192	59	-0.16	600.98	0
600	Total Aflatoxin (ppb)	0354	11.20	0.7400	11.75	4.046	1.192	59	-0.14	600.22	0
600	Total Aflatoxin (ppb)	0723	11.24	0.0000	11.75	4.046	1.192	59	-0.13	600.20	0
600	Total Aflatoxin (ppb)	0034	11.65	0.8000	11.75	4.046	1.192	59	-0.02	600.01	0
600	Total Aflatoxin (ppb)	0656	11.65	0.0000	11.75	4.046	1.192	59	-0.02	600.12	0
600	Total Aflatoxin (ppb)	2113	11.75	0.7000	11.75	4.046	1.192	59	0.00	600.01	0
600	Total Aflatoxin (ppb)	0227	12.00	4.000	11.75	4.046	1.192	59	0.06	600.01	0
600	Total Aflatoxin (ppb)	2196	12.00	0.0000	11.75	4.046	1.192	59	0.06	600.08	0
600	Total Aflatoxin (ppb)	0563	12.01	0.1100	11.75	4.046	1.192	59	0.06	600.20	0

Test Material Code # 202564			Analyte Proficiency Testing Report						Issue Date : 11/30/2025		
Method	Analyte	Lab	Lab Data		Robust* Analyte Values				AAFCO PT	Your	
Group	Group (Units)	Code	Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests	Z Score	Method	Flag
600	Total Aflatoxin (ppb)	2096	12.05	1.500	11.75	4.046	1.192	59	0.08	600.01	0
600	Total Aflatoxin (ppb)	0629	12.31	1.120	11.75	4.046	1.192	59	0.14	600.20	0
600	Total Aflatoxin (ppb)	0520	12.50	3.000	11.75	4.046	1.192	59	0.19	600.14	0
600	Total Aflatoxin (ppb)	0353	12.62	0.8310	11.75	4.046	1.192	59	0.22	600.24	0
600	Total Aflatoxin (ppb)	0001	12.69	0.0500	11.75	4.046	1.192	59	0.23	600.20	0
600	Total Aflatoxin (ppb)	0035	13.05	0.5000	11.75	4.046	1.192	59	0.32	600.12	0
600	Total Aflatoxin (ppb)	2327	13.05	1.700	11.75	4.046	1.192	59	0.32	600.24	0
600	Total Aflatoxin (ppb)	2352	13.34	1.330	11.75	4.046	1.192	59	0.39	600.23	0
600	Total Aflatoxin (ppb)	0089	13.35	2.500	11.75	4.046	1.192	59	0.40	600.02	0
600	Total Aflatoxin (ppb)	0407	13.40	6.560	11.75	4.046	1.192	59	0.41	600.13	0
600	Total Aflatoxin (ppb)	0529	14.10	3.000	11.75	4.046	1.192	59	0.58	600.01	0
600	Total Aflatoxin (ppb)	2022	14.15	0.9200	11.75	4.046	1.192	59	0.59	600.01	0
600	Total Aflatoxin (ppb)	0098	14.20	1.510	11.75	4.046	1.192	59	0.61	600.22	0
600	Total Aflatoxin (ppb)	2022	14.21	0.6300	11.75	4.046	1.192	59	0.61	600.08	0
600	Total Aflatoxin (ppb)	0354	14.50	7.710	11.75	4.046	1.192	59	0.68	600.08	0
600	Total Aflatoxin (ppb)	2456	14.54	2.814	11.75	4.046	1.192	59	0.69	600.99	0
600	Total Aflatoxin (ppb)	0035	15.15	1.300	11.75	4.046	1.192	59	0.84	600.21	0
600	Total Aflatoxin (ppb)	0590	15.55	1.900	11.75	4.046	1.192	59	0.94	600.12	0
600	Total Aflatoxin (ppb)	0042	15.82	2.510	11.75	4.046	1.192	59	1.01	600.20	0
600	Total Aflatoxin (ppb)	0148	15.84	1.670	11.75	4.046	1.192	59	1.01	600.24	0
600	Total Aflatoxin (ppb)	2239	17.58	1.850	11.75	4.046	1.192	59	1.44	600.98	0
600	Total Aflatoxin (ppb)	0218	17.80	0.2000	11.75	4.046	1.192	59	1.50	600.24	0
600	Total Aflatoxin (ppb)	2239	19.08	1.300	11.75	4.046	1.192	59	1.81	600.20	0
600	Total Aflatoxin (ppb)	2129	20.75	2.700	11.75	4.046	1.192	59	2.23	600.24	0
600	Total Aflatoxin (ppb)	0596	27.15	0.9000	11.75	4.046	1.192	59	3.81	600.98	0
600	Total Aflatoxin (ppb)	2302	28.50	0.6500	11.75	4.046	1.192	59	4.14	600.99	0
601	Aflatoxin B1 (ppb)	2326	4.500	1.000	11.44	3.945	1.002	37	-1.76	601.23	0
601	Aflatoxin B1 (ppb)	0553	5.230	3.200	11.44	3.945	1.002	37	-1.58	601.23	0
601	Aflatoxin B1 (ppb)	2192	5.650	0.1000	11.44	3.945	1.002	37	-1.47	601.20	0
601	Aflatoxin B1 (ppb)	2369	6.755	0.2900	11.44	3.945	1.002	37	-1.19	601.20	0
601	Aflatoxin B1 (ppb)	2369	6.755	0.2900	11.44	3.945	1.002	37	-1.19	601.24	0
601	Aflatoxin B1 (ppb)	0675	7.035	0.0700	11.44	3.945	1.002	37	-1.12	601.20	0
601	Aflatoxin B1 (ppb)	2196	7.690	0.0000	11.44	3.945	1.002	37	-0.95	601.20	0
601	Aflatoxin B1 (ppb)	2066	7.755	0.0300	11.44	3.945	1.002	37	-0.94	601.20	0
601	Aflatoxin B1 (ppb)	0964	8.000	0.0000	11.44	3.945	1.002	37	-0.87	601.23	0
601	Aflatoxin B1 (ppb)	0682	8.500	0.0000	11.44	3.945	1.002	37	-0.75	601.23	0
601	Aflatoxin B1 (ppb)	0010	8.900	0.8000	11.44	3.945	1.002	37	-0.65	601.20	0
601	Aflatoxin B1 (ppb)	0202	9.440	0.4800	11.44	3.945	1.002	37	-0.51	601.23	0
601	Aflatoxin B1 (ppb)	2033	9.750	2.500	11.44	3.945	1.002	37	-0.43	601.23	0
601	Aflatoxin B1 (ppb)	0870	10.09	0.1250	11.44	3.945	1.002	37	-0.34	601.23	0
601	Aflatoxin B1 (ppb)	0354	10.32	0.7200	11.44	3.945	1.002	37	-0.29	601.24	0
601	Aflatoxin B1 (ppb)	0629	11.43	1.260	11.44	3.945	1.002	37	0.00	601.20	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT Z Score	Your	
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests		Method	Flag
601	Aflatoxin B1 (ppb)	2022	11.45	0.2300	11.44	3.945	1.002	37	0.00	601.98	0
601	Aflatoxin B1 (ppb)	0028	11.52	0.0400	11.44	3.945	1.002	37	0.02	601.23	0
601	Aflatoxin B1 (ppb)	2352	11.72	1.220	11.44	3.945	1.002	37	0.07	601.23	0
601	Aflatoxin B1 (ppb)	0563	12.01	0.1100	11.44	3.945	1.002	37	0.14	601.20	0
601	Aflatoxin B1 (ppb)	0353	12.09	1.290	11.44	3.945	1.002	37	0.16	601.23	0
601	Aflatoxin B1 (ppb)	2327	12.15	1.300	11.44	3.945	1.002	37	0.18	601.23	0
601	Aflatoxin B1 (ppb)	0098	12.88	1.750	11.44	3.945	1.002	37	0.36	601.24	0
601	Aflatoxin B1 (ppb)	2022	13.30	0.5000	11.44	3.945	1.002	37	0.47	601.16	0
601	Aflatoxin B1 (ppb)	2055	13.40	1.000	11.44	3.945	1.002	37	0.50	601.23	0
601	Aflatoxin B1 (ppb)	0910	13.49	1.880	11.44	3.945	1.002	37	0.52	601.21	0
601	Aflatoxin B1 (ppb)	2022	13.65	0.8500	11.44	3.945	1.002	37	0.56	601.91	0
601	Aflatoxin B1 (ppb)	0913	13.70	3.000	11.44	3.945	1.002	37	0.57	601.23	0
601	Aflatoxin B1 (ppb)	0148	13.77	0.0100	11.44	3.945	1.002	37	0.59	601.23	0
601	Aflatoxin B1 (ppb)	0297	14.15	5.300	11.44	3.945	1.002	37	0.69	601.23	0
601	Aflatoxin B1 (ppb)	0035	14.20	1.200	11.44	3.945	1.002	37	0.70	601.21	0
601	Aflatoxin B1 (ppb)	0009	14.95	1.952	11.44	3.945	1.002	37	0.89	601.23	0
601	Aflatoxin B1 (ppb)	0218	16.55	0.1000	11.44	3.945	1.002	37	1.29	601.23	0
601	Aflatoxin B1 (ppb)	2239	17.12	0.1600	11.44	3.945	1.002	37	1.44	601.98	0
601	Aflatoxin B1 (ppb)	0505	17.87	1.870	11.44	3.945	1.002	37	1.63	601.16	0
601	Aflatoxin B1 (ppb)	2239	18.32	1.240	11.44	3.945	1.002	37	1.74	601.20	0
601	Aflatoxin B1 (ppb)	2140	20.00	4.000	11.44	3.945	1.002	37	2.17	601.23	0
601	Aflatoxin B1 (ppb)	0242	< 1		11.44	3.945	1.002	37		601.23	5
602	Aflatoxin B2 (ppb)	0675	0.4200	0.0600	0.8712	0.3220	0.1012	19	-1.40	602.20	0
602	Aflatoxin B2 (ppb)	2192	0.5250	0.0300	0.8712	0.3220	0.1012	19	-1.08	602.20	0
602	Aflatoxin B2 (ppb)	0682	0.6000	0.0000	0.8712	0.3220	0.1012	19	-0.84	602.23	0
602	Aflatoxin B2 (ppb)	2196	0.6300	0.0000	0.8712	0.3220	0.1012	19	-0.75	602.20	0
602	Aflatoxin B2 (ppb)	2369	0.6300	0.0000	0.8712	0.3220	0.1012	19	-0.75	602.20	0
602	Aflatoxin B2 (ppb)	2369	0.6300	0.0000	0.8712	0.3220	0.1012	19	-0.75	602.24	0
602	Aflatoxin B2 (ppb)	0870	0.6750	0.1000	0.8712	0.3220	0.1012	19	-0.61	602.23	0
602	Aflatoxin B2 (ppb)	0353	0.7190	0.0860	0.8712	0.3220	0.1012	19	-0.47	602.23	0
602	Aflatoxin B2 (ppb)	0910	0.8700	0.0600	0.8712	0.3220	0.1012	19	0.00	602.21	0
602	Aflatoxin B2 (ppb)	0629	0.8800	0.1400	0.8712	0.3220	0.1012	19	0.03	602.20	0
602	Aflatoxin B2 (ppb)	0354	0.8815	0.0090	0.8712	0.3220	0.1012	19	0.03	602.24	0
602	Aflatoxin B2 (ppb)	2327	0.9000	0.4000	0.8712	0.3220	0.1012	19	0.09	602.23	0
602	Aflatoxin B2 (ppb)	0035	0.9200	0.1200	0.8712	0.3220	0.1012	19	0.15	602.21	0
602	Aflatoxin B2 (ppb)	0964	1.000	0.0000	0.8712	0.3220	0.1012	19	0.40	602.23	0
602	Aflatoxin B2 (ppb)	0913	1.150	0.1000	0.8712	0.3220	0.1012	19	0.87	602.23	0
602	Aflatoxin B2 (ppb)	2352	1.205	0.1300	0.8712	0.3220	0.1012	19	1.04	602.23	0
602	Aflatoxin B2 (ppb)	0218	1.250	0.1000	0.8712	0.3220	0.1012	19	1.18	602.23	0
602	Aflatoxin B2 (ppb)	0098	1.320	0.2400	0.8712	0.3220	0.1012	19	1.39	602.24	0
602	Aflatoxin B2 (ppb)	0010	1.750	0.1000	0.8712	0.3220	0.1012	19	2.73	602.20	0
602	Aflatoxin B2 (ppb)	0028	0.0000	0.0000	0.8712	0.3220	0.1012	19		602.23	4

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests			
602	Aflatoxin B2 (ppb)	0297	0.0000	0.0000	0.8712	0.3220	0.1012	19		602.23	4
602	Aflatoxin B2 (ppb)	2326	0.0000	0.0000	0.8712	0.3220	0.1012	19		602.23	4
602	Aflatoxin B2 (ppb)	2066	0.0150	0.0300	0.8712	0.3220	0.1012	19		602.20	4
602	Aflatoxin B2 (ppb)	0009	< 0.68		0.8712	0.3220	0.1012	19		602.23	5
602	Aflatoxin B2 (ppb)	0202	< 1		0.8712	0.3220	0.1012	19		602.23	5
602	Aflatoxin B2 (ppb)	0242	< 1		0.8712	0.3220	0.1012	19		602.23	5
602	Aflatoxin B2 (ppb)	2033	< 1		0.8712	0.3220	0.1012	19		602.23	5
602	Aflatoxin B2 (ppb)	0148	< 2.5		0.8712	0.3220	0.1012	19		602.23	5
602	Aflatoxin B2 (ppb)	0563	< 5		0.8712	0.3220	0.1012	19		602.20	5
602	Aflatoxin B2 (ppb)	2140	< 20		0.8712	0.3220	0.1012	19		602.23	5
603	Aflatoxin G1 (ppb)	0910	0.1200	0.0400				2		603.21	0
603	Aflatoxin G1 (ppb)	0505	2.110	0.0200				2		603.16	0
603	Aflatoxin G1 (ppb)	2066	0.0000	0.0000				2		603.20	4
603	Aflatoxin G1 (ppb)	0028	0.0000	0.0000				2		603.23	4
603	Aflatoxin G1 (ppb)	2326	0.0000	0.0000				2		603.23	4
603	Aflatoxin G1 (ppb)	0629	< 0.25					2		603.20	5
603	Aflatoxin G1 (ppb)	0682	< 0.5					2		603.23	5
603	Aflatoxin G1 (ppb)	2352	< 0.5					2		603.23	5
603	Aflatoxin G1 (ppb)	0870	< 0.6					2		603.23	5
603	Aflatoxin G1 (ppb)	2327	< 0.6					2		603.23	5
603	Aflatoxin G1 (ppb)	0009	< 0.68					2		603.23	5
603	Aflatoxin G1 (ppb)	0202	< 1					2		603.23	5
603	Aflatoxin G1 (ppb)	0218	< 1					2		603.23	5
603	Aflatoxin G1 (ppb)	0242	< 1					2		603.23	5
603	Aflatoxin G1 (ppb)	0913	< 1					2		603.23	5
603	Aflatoxin G1 (ppb)	0964	< 1					2		603.23	5
603	Aflatoxin G1 (ppb)	2033	< 1					2		603.23	5
603	Aflatoxin G1 (ppb)	0148	< 2.5					2		603.23	5
603	Aflatoxin G1 (ppb)	0098	< 15					2		603.24	5
603	Aflatoxin G1 (ppb)	2140	< 20					2		603.23	5
604	Aflatoxin G2 (ppb)	2066	0.0000	0.0000				0		604.20	4
604	Aflatoxin G2 (ppb)	0028	0.0000	0.0000				0		604.23	4
604	Aflatoxin G2 (ppb)	2326	0.0000	0.0000				0		604.23	4
604	Aflatoxin G2 (ppb)	0910	< 0.01					0		604.21	5
604	Aflatoxin G2 (ppb)	0629	< 0.25					0		604.20	5
604	Aflatoxin G2 (ppb)	0682	< 0.5					0		604.23	5
604	Aflatoxin G2 (ppb)	2327	< 0.6					0		604.23	5
604	Aflatoxin G2 (ppb)	0009	< 0.69					0		604.23	5
604	Aflatoxin G2 (ppb)	0202	< 1					0		604.23	5
604	Aflatoxin G2 (ppb)	0218	< 1					0		604.23	5
604	Aflatoxin G2 (ppb)	0242	< 1					0		604.23	5
604	Aflatoxin G2 (ppb)	0870	< 1					0		604.23	5

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests			
604	Aflatoxin G2 (ppb)	0913	< 1					0		604.23	5
604	Aflatoxin G2 (ppb)	0964	< 1					0		604.23	5
604	Aflatoxin G2 (ppb)	2033	< 1					0		604.23	5
604	Aflatoxin G2 (ppb)	0148	< 3					0		604.23	5
604	Aflatoxin G2 (ppb)	0098	< 5					0		604.24	5
604	Aflatoxin G2 (ppb)	2140	< 20					0		604.23	5
610	Deoxynivalenol (ppb)	0553	565.9	38.20	1,670	503.0	105.8	56	-2.20	610.23	0
610	Deoxynivalenol (ppb)	2066	800.0	60.00	1,670	503.0	105.8	56	-1.73	610.07	0
610	Deoxynivalenol (ppb)	0964	938.0	50.00	1,670	503.0	105.8	56	-1.46	610.23	0
610	Deoxynivalenol (ppb)	0242	1,019	130.0	1,670	503.0	105.8	56	-1.29	610.23	0
610	Deoxynivalenol (ppb)	2327	1,170	74.30	1,670	503.0	105.8	56	-1.00	610.23	0
610	Deoxynivalenol (ppb)	2268	1,230	37.00	1,670	503.0	105.8	56	-0.88	610.22	0
610	Deoxynivalenol (ppb)	2140	1,250	302.0	1,670	503.0	105.8	56	-0.84	610.23	0
610	Deoxynivalenol (ppb)	2129	1,263	12.00	1,670	503.0	105.8	56	-0.81	610.23	0
610	Deoxynivalenol (ppb)	0028	1,296	120.1	1,670	503.0	105.8	56	-0.74	610.23	0
610	Deoxynivalenol (ppb)	0278	1,350	100.0	1,670	503.0	105.8	56	-0.64	610.09	0
610	Deoxynivalenol (ppb)	2033	1,400	0.0000	1,670	503.0	105.8	56	-0.54	610.23	0
610	Deoxynivalenol (ppb)	0910	1,435	170.0	1,670	503.0	105.8	56	-0.47	610.20	0
610	Deoxynivalenol (ppb)	0511	1,440	80.00	1,670	503.0	105.8	56	-0.46	610.98	0
610	Deoxynivalenol (ppb)	0407	1,445	148.0	1,670	503.0	105.8	56	-0.45	610.07	0
610	Deoxynivalenol (ppb)	2352	1,516	87.36	1,670	503.0	105.8	56	-0.31	610.23	0
610	Deoxynivalenol (ppb)	0148	1,528	22.21	1,670	503.0	105.8	56	-0.28	610.23	0
610	Deoxynivalenol (ppb)	0098	1,550	100.0	1,670	503.0	105.8	56	-0.24	610.20	0
610	Deoxynivalenol (ppb)	2055	1,556	67.00	1,670	503.0	105.8	56	-0.23	610.23	0
610	Deoxynivalenol (ppb)	0010	1,580	71.00	1,670	503.0	105.8	56	-0.18	610.23	0
610	Deoxynivalenol (ppb)	0656	1,595	0.0000	1,670	503.0	105.8	56	-0.15	610.07	0
610	Deoxynivalenol (ppb)	0353	1,605	365.0	1,670	503.0	105.8	56	-0.13	610.23	0
610	Deoxynivalenol (ppb)	0009	1,637	28.87	1,670	503.0	105.8	56	-0.07	610.23	0
610	Deoxynivalenol (ppb)	0297	1,638	75.00	1,670	503.0	105.8	56	-0.07	610.23	0
610	Deoxynivalenol (ppb)	0529	1,650	500.0	1,670	503.0	105.8	56	-0.04	610.01	0
610	Deoxynivalenol (ppb)	0674	1,650	0.0000	1,670	503.0	105.8	56	-0.04	610.20	0
610	Deoxynivalenol (ppb)	0913	1,650	100.0	1,670	503.0	105.8	56	-0.04	610.23	0
610	Deoxynivalenol (ppb)	0218	1,680	140.0	1,670	503.0	105.8	56	0.02	610.23	0
610	Deoxynivalenol (ppb)	0682	1,680	0.0000	1,670	503.0	105.8	56	0.02	610.23	0
610	Deoxynivalenol (ppb)	2326	1,690	153.0	1,670	503.0	105.8	56	0.04	610.23	0
610	Deoxynivalenol (ppb)	0089	1,693	411.0	1,670	503.0	105.8	56	0.04	610.02	0
610	Deoxynivalenol (ppb)	2113	1,700	0.0000	1,670	503.0	105.8	56	0.06	610.01	0
610	Deoxynivalenol (ppb)	2239	1,701	2.000	1,670	503.0	105.8	56	0.06	610.20	0
610	Deoxynivalenol (ppb)	0629	1,715	50.00	1,670	503.0	105.8	56	0.09	610.07	0
610	Deoxynivalenol (ppb)	0870	1,720	0.0000	1,670	503.0	105.8	56	0.10	610.99	0
610	Deoxynivalenol (ppb)	2196	1,736	0.0000	1,670	503.0	105.8	56	0.13	610.20	0
610	Deoxynivalenol (ppb)	0590	1,745	50.00	1,670	503.0	105.8	56	0.15	610.07	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests			
610	Deoxynivalenol (ppb)	2302	1,785	76.83	1,670	503.0	105.8	56	0.23	610.99	0
610	Deoxynivalenol (ppb)	2022	1,790	160.0	1,670	503.0	105.8	56	0.24	610.98	0
610	Deoxynivalenol (ppb)	0227	1,800	200.0	1,670	503.0	105.8	56	0.26	610.01	0
610	Deoxynivalenol (ppb)	2022	1,815	30.00	1,670	503.0	105.8	56	0.29	610.07	0
610	Deoxynivalenol (ppb)	2239	1,843	61.00	1,670	503.0	105.8	56	0.34	610.98	0
610	Deoxynivalenol (ppb)	0042	1,850	100.0	1,670	503.0	105.8	56	0.36	610.01	0
610	Deoxynivalenol (ppb)	0208	1,850	140.0	1,670	503.0	105.8	56	0.36	610.01	0
610	Deoxynivalenol (ppb)	0554	1,900	200.0	1,670	503.0	105.8	56	0.46	610.01	0
610	Deoxynivalenol (ppb)	0505	1,925	8.580	1,670	503.0	105.8	56	0.51	610.16	0
610	Deoxynivalenol (ppb)	0202	1,931	65.00	1,670	503.0	105.8	56	0.52	610.23	0
610	Deoxynivalenol (ppb)	0035	1,942	16.00	1,670	503.0	105.8	56	0.54	610.08	0
610	Deoxynivalenol (ppb)	0870	1,960	80.00	1,670	503.0	105.8	56	0.58	610.23	0
610	Deoxynivalenol (ppb)	0034	1,975	50.00	1,670	503.0	105.8	56	0.61	610.01	0
610	Deoxynivalenol (ppb)	0354	2,041	312.6	1,670	503.0	105.8	56	0.74	610.05	0
610	Deoxynivalenol (ppb)	2022	2,110	40.00	1,670	503.0	105.8	56	0.87	610.01	0
610	Deoxynivalenol (ppb)	2196	2,171	0.0000	1,670	503.0	105.8	56	1.00	610.05	0
610	Deoxynivalenol (ppb)	0066	2,200	600.0	1,670	503.0	105.8	56	1.05	610.08	0
610	Deoxynivalenol (ppb)	2022	2,225	10.00	1,670	503.0	105.8	56	1.10	610.05	0
610	Deoxynivalenol (ppb)	0563	2,239	283.0	1,670	503.0	105.8	56	1.13	610.25	0
610	Deoxynivalenol (ppb)	0596	2,250	300.0	1,670	503.0	105.8	56	1.15	610.98	0
610	Deoxynivalenol (ppb)	0027	1,196	2,388	1,670	503.0	105.8	56	-0.94	610.01	1
620	Total Fumonisin (ppb)	2369	807.9	14.78	1,582	477.0	145.4	38	-1.62	620.11	0
620	Total Fumonisin (ppb)	0520	950.0	100.0	1,582	477.0	145.4	38	-1.32	620.12	0
620	Total Fumonisin (ppb)	2302	1,006	199.3	1,582	477.0	145.4	38	-1.21	620.99	0
620	Total Fumonisin (ppb)	0407	1,095	90.00	1,582	477.0	145.4	38	-1.02	620.09	0
620	Total Fumonisin (ppb)	0511	1,095	190.0	1,582	477.0	145.4	38	-1.02	620.98	0
620	Total Fumonisin (ppb)	0035	1,185	10.00	1,582	477.0	145.4	38	-0.83	620.09	0
620	Total Fumonisin (ppb)	0098	1,200	200.0	1,582	477.0	145.4	38	-0.80	620.11	0
620	Total Fumonisin (ppb)	2066	1,240	120.0	1,582	477.0	145.4	38	-0.72	620.09	0
620	Total Fumonisin (ppb)	2192	1,300	0.0000	1,582	477.0	145.4	38	-0.59	620.01	0
620	Total Fumonisin (ppb)	2352	1,303	57.50	1,582	477.0	145.4	38	-0.58	620.23	0
620	Total Fumonisin (ppb)	0505	1,323	11.79	1,582	477.0	145.4	38	-0.54	620.16	0
620	Total Fumonisin (ppb)	0674	1,330	0.0000	1,582	477.0	145.4	38	-0.53	620.07	0
620	Total Fumonisin (ppb)	0010	1,339	99.00	1,582	477.0	145.4	38	-0.51	620.23	0
620	Total Fumonisin (ppb)	0682	1,340	0.0000	1,582	477.0	145.4	38	-0.51	620.23	0
620	Total Fumonisin (ppb)	0034	1,353	145.0	1,582	477.0	145.4	38	-0.48	620.01	0
620	Total Fumonisin (ppb)	2033	1,400	200.0	1,582	477.0	145.4	38	-0.38	620.23	0
620	Total Fumonisin (ppb)	0656	1,424	0.0000	1,582	477.0	145.4	38	-0.33	620.09	0
620	Total Fumonisin (ppb)	0027	1,425	230.0	1,582	477.0	145.4	38	-0.33	620.01	0
620	Total Fumonisin (ppb)	0042	1,450	300.0	1,582	477.0	145.4	38	-0.28	620.01	0
620	Total Fumonisin (ppb)	2239	1,505	30.00	1,582	477.0	145.4	38	-0.16	620.98	0
620	Total Fumonisin (ppb)	2239	1,588	48.00	1,582	477.0	145.4	38	0.01	620.20	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT Z Score	Your	
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests		Method	Flag
620	Total Fumonisin (ppb)	2268	1,650	100.0	1,582	477.0	145.4	38	0.14	620.11	0
620	Total Fumonisin (ppb)	2113	1,700	400.0	1,582	477.0	145.4	38	0.25	620.01	0
620	Total Fumonisin (ppb)	2022	1,750	100.0	1,582	477.0	145.4	38	0.35	620.08	0
620	Total Fumonisin (ppb)	0202	1,753	64.00	1,582	477.0	145.4	38	0.36	620.23	0
620	Total Fumonisin (ppb)	2327	1,821	137.1	1,582	477.0	145.4	38	0.50	620.23	0
620	Total Fumonisin (ppb)	2129	1,867	143.0	1,582	477.0	145.4	38	0.60	620.23	0
620	Total Fumonisin (ppb)	2022	1,935	90.00	1,582	477.0	145.4	38	0.74	620.01	0
620	Total Fumonisin (ppb)	0554	1,950	1,500	1,582	477.0	145.4	38	0.77	620.01	0
620	Total Fumonisin (ppb)	2022	1,950	120.0	1,582	477.0	145.4	38	0.77	620.02	0
620	Total Fumonisin (ppb)	0353	1,977	101.0	1,582	477.0	145.4	38	0.83	620.23	0
620	Total Fumonisin (ppb)	0227	2,000	0.0000	1,582	477.0	145.4	38	0.88	620.07	0
620	Total Fumonisin (ppb)	0089	2,047	239.0	1,582	477.0	145.4	38	0.97	620.03	0
620	Total Fumonisin (ppb)	0066	2,100	200.0	1,582	477.0	145.4	38	1.09	620.09	0
620	Total Fumonisin (ppb)	0208	2,100	200.0	1,582	477.0	145.4	38	1.09	620.98	0
620	Total Fumonisin (ppb)	2022	2,265	50.00	1,582	477.0	145.4	38	1.43	620.07	0
620	Total Fumonisin (ppb)	0529	2,300	400.0	1,582	477.0	145.4	38	1.51	620.02	0
620	Total Fumonisin (ppb)	0596	3,300	800.0	1,582	477.0	145.4	38	3.60	620.98	0
621	Fumonsin B1 (ppb)	0009	684.1	50.53	997.8	304.7	83.88	23	-1.03	621.23	0
621	Fumonsin B1 (ppb)	0297	685.0	70.00	997.8	304.7	83.88	23	-1.03	621.23	0
621	Fumonsin B1 (ppb)	0913	700.0	0.0000	997.8	304.7	83.88	23	-0.98	621.23	0
621	Fumonsin B1 (ppb)	2140	707.0	266.0	997.8	304.7	83.88	23	-0.95	621.23	0
621	Fumonsin B1 (ppb)	0563	757.0	59.51	997.8	304.7	83.88	23	-0.79	621.20	0
621	Fumonsin B1 (ppb)	2369	807.9	14.78	997.8	304.7	83.88	23	-0.62	621.21	0
621	Fumonsin B1 (ppb)	0218	896.0	166.0	997.8	304.7	83.88	23	-0.33	621.23	0
621	Fumonsin B1 (ppb)	0098	900.0	200.0	997.8	304.7	83.88	23	-0.32	621.21	0
621	Fumonsin B1 (ppb)	0964	928.5	69.00	997.8	304.7	83.88	23	-0.23	621.23	0
621	Fumonsin B1 (ppb)	2352	930.4	48.21	997.8	304.7	83.88	23	-0.22	621.23	0
621	Fumonsin B1 (ppb)	0148	949.6	35.55	997.8	304.7	83.88	23	-0.16	621.23	0
621	Fumonsin B1 (ppb)	0010	951.5	5.000	997.8	304.7	83.88	23	-0.15	621.23	0
621	Fumonsin B1 (ppb)	0910	980.0	0.0000	997.8	304.7	83.88	23	-0.06	621.21	0
621	Fumonsin B1 (ppb)	0028	1,000	177.3	997.8	304.7	83.88	23	0.01	621.23	0
621	Fumonsin B1 (ppb)	0682	1,040	0.0000	997.8	304.7	83.88	23	0.14	621.23	0
621	Fumonsin B1 (ppb)	0870	1,040	40.00	997.8	304.7	83.88	23	0.14	621.23	0
621	Fumonsin B1 (ppb)	2055	1,098	87.00	997.8	304.7	83.88	23	0.33	621.23	0
621	Fumonsin B1 (ppb)	2033	1,100	200.0	997.8	304.7	83.88	23	0.34	621.23	0
621	Fumonsin B1 (ppb)	0202	1,302	48.00	997.8	304.7	83.88	23	1.00	621.23	0
621	Fumonsin B1 (ppb)	2327	1,316	83.10	997.8	304.7	83.88	23	1.04	621.23	0
621	Fumonsin B1 (ppb)	2326	1,505	75.00	997.8	304.7	83.88	23	1.66	621.23	0
621	Fumonsin B1 (ppb)	0353	1,560	28.00	997.8	304.7	83.88	23	1.85	621.23	0
621	Fumonsin B1 (ppb)	0553	1,879	53.00	997.8	304.7	83.88	23	2.89	621.23	0
622	Fumonisin B2 (ppb)	2033	195.0	10.00	276.4	87.40	27.19	21	-0.93	622.23	0
622	Fumonisin B2 (ppb)	0913	200.0	0.0000	276.4	87.40	27.19	21	-0.87	622.23	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests			
622	Fumonisin B2 (ppb)	0009	210.2	24.53	276.4	87.40	27.19	21	-0.76	622.23	0
622	Fumonisin B2 (ppb)	0028	210.6	22.32	276.4	87.40	27.19	21	-0.75	622.23	0
622	Fumonisin B2 (ppb)	0218	227.5	21.00	276.4	87.40	27.19	21	-0.56	622.23	0
622	Fumonisin B2 (ppb)	0964	245.0	18.00	276.4	87.40	27.19	21	-0.36	622.23	0
622	Fumonisin B2 (ppb)	2326	245.0	2.000	276.4	87.40	27.19	21	-0.36	622.23	0
622	Fumonisin B2 (ppb)	2352	246.1	6.360	276.4	87.40	27.19	21	-0.35	622.23	0
622	Fumonisin B2 (ppb)	0910	254.0	0.0000	276.4	87.40	27.19	21	-0.26	622.21	0
622	Fumonisin B2 (ppb)	0010	268.0	62.00	276.4	87.40	27.19	21	-0.10	622.23	0
622	Fumonisin B2 (ppb)	0553	276.1	30.90	276.4	87.40	27.19	21	0.00	622.23	0
622	Fumonisin B2 (ppb)	0353	287.5	17.00	276.4	87.40	27.19	21	0.13	622.23	0
622	Fumonisin B2 (ppb)	0870	290.0	20.00	276.4	87.40	27.19	21	0.16	622.23	0
622	Fumonisin B2 (ppb)	0098	300.0	0.0000	276.4	87.40	27.19	21	0.27	622.21	0
622	Fumonisin B2 (ppb)	0682	300.0	0.0000	276.4	87.40	27.19	21	0.27	622.23	0
622	Fumonisin B2 (ppb)	0148	301.1	8.290	276.4	87.40	27.19	21	0.28	622.23	0
622	Fumonisin B2 (ppb)	0202	331.5	1.000	276.4	87.40	27.19	21	0.63	622.23	0
622	Fumonisin B2 (ppb)	2055	335.5	37.00	276.4	87.40	27.19	21	0.68	622.23	0
622	Fumonisin B2 (ppb)	2140	347.0	202.0	276.4	87.40	27.19	21	0.81	622.23	0
622	Fumonisin B2 (ppb)	2327	424.8	54.00	276.4	87.40	27.19	21	1.70	622.23	0
622	Fumonisin B2 (ppb)	0297	480.0	160.0	276.4	87.40	27.19	21	2.33	622.23	0
623	Fumonisin B3 (ppb)	0353	76.50	15.00	114.3	37.02	15.12	10	-1.02	623.23	0
623	Fumonisin B3 (ppb)	0009	86.36	10.82	114.3	37.02	15.12	10	-0.75	623.23	0
623	Fumonisin B3 (ppb)	2033	98.00	4.000	114.3	37.02	15.12	10	-0.44	623.23	0
623	Fumonisin B3 (ppb)	0913	100.0	0.0000	114.3	37.02	15.12	10	-0.39	623.23	0
623	Fumonisin B3 (ppb)	0010	119.0	42.00	114.3	37.02	15.12	10	0.13	623.23	0
623	Fumonisin B3 (ppb)	0202	119.5	15.00	114.3	37.02	15.12	10	0.14	623.23	0
623	Fumonisin B3 (ppb)	0870	120.0	0.0000	114.3	37.02	15.12	10	0.15	623.23	0
623	Fumonisin B3 (ppb)	2352	126.9	2.940	114.3	37.02	15.12	10	0.34	623.23	0
623	Fumonisin B3 (ppb)	2326	141.0	8.000	114.3	37.02	15.12	10	0.72	623.23	0
623	Fumonisin B3 (ppb)	0553	335.6	30.10	114.3	37.02	15.12	10	5.98	623.23	0
623	Fumonisin B3 (ppb)	2327	< 160		114.3	37.02	15.12	10		623.23	5
623	Fumonisin B3 (ppb)	0964	< 200		114.3	37.02	15.12	10		623.23	5
630	Ochratoxin A (ppb)	0227	22.00	4.000	55.58	18.35	4.470	44	-1.83	630.01	0
630	Ochratoxin A (ppb)	2326	31.00	0.0000	55.58	18.35	4.470	44	-1.34	630.23	0
630	Ochratoxin A (ppb)	0870	31.44	0.6400	55.58	18.35	4.470	44	-1.32	630.23	0
630	Ochratoxin A (ppb)	0354	31.72	6.680	55.58	18.35	4.470	44	-1.30	630.03	0
630	Ochratoxin A (ppb)	0964	33.00	2.000	55.58	18.35	4.470	44	-1.23	630.23	0
630	Ochratoxin A (ppb)	0353	38.50	5.000	55.58	18.35	4.470	44	-0.93	630.23	0
630	Ochratoxin A (ppb)	0629	39.00	10.00	55.58	18.35	4.470	44	-0.90	630.20	0
630	Ochratoxin A (ppb)	0675	40.14	3.140	55.58	18.35	4.470	44	-0.84	630.20	0
630	Ochratoxin A (ppb)	0520	41.50	3.000	55.58	18.35	4.470	44	-0.77	630.06	0
630	Ochratoxin A (ppb)	0682	43.10	0.0000	55.58	18.35	4.470	44	-0.68	630.23	0
630	Ochratoxin A (ppb)	0723	43.39	0.0000	55.58	18.35	4.470	44	-0.66	630.20	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests			
630	Ochratoxin A (ppb)	2302	45.23	8.630	55.58	18.35	4.470	44	-0.56	630.99	0
630	Ochratoxin A (ppb)	2022	45.31	2.630	55.58	18.35	4.470	44	-0.56	630.01	0
630	Ochratoxin A (ppb)	0505	47.57	10.78	55.58	18.35	4.470	44	-0.44	630.16	0
630	Ochratoxin A (ppb)	0596	49.70	3.800	55.58	18.35	4.470	44	-0.32	630.98	0
630	Ochratoxin A (ppb)	0910	50.20	0.0000	55.58	18.35	4.470	44	-0.29	630.21	0
630	Ochratoxin A (ppb)	2239	51.10	2.800	55.58	18.35	4.470	44	-0.24	630.98	0
630	Ochratoxin A (ppb)	0202	52.90	1.400	55.58	18.35	4.470	44	-0.15	630.23	0
630	Ochratoxin A (ppb)	0066	52.95	5.900	55.58	18.35	4.470	44	-0.14	630.05	0
630	Ochratoxin A (ppb)	2369	53.84	2.570	55.58	18.35	4.470	44	-0.10	630.06	0
630	Ochratoxin A (ppb)	2369	53.84	2.570	55.58	18.35	4.470	44	-0.10	630.20	0
630	Ochratoxin A (ppb)	2022	54.17	2.490	55.58	18.35	4.470	44	-0.08	630.98	0
630	Ochratoxin A (ppb)	0354	54.81	2.827	55.58	18.35	4.470	44	-0.04	630.20	0
630	Ochratoxin A (ppb)	2022	54.87	2.220	55.58	18.35	4.470	44	-0.04	630.04	0
630	Ochratoxin A (ppb)	2033	57.50	13.00	55.58	18.35	4.470	44	0.10	630.23	0
630	Ochratoxin A (ppb)	2066	57.57	0.3700	55.58	18.35	4.470	44	0.11	630.20	0
630	Ochratoxin A (ppb)	0656	58.42	0.0000	55.58	18.35	4.470	44	0.15	630.05	0
630	Ochratoxin A (ppb)	0407	59.70	2.140	55.58	18.35	4.470	44	0.22	630.04	0
630	Ochratoxin A (ppb)	0010	60.70	2.800	55.58	18.35	4.470	44	0.28	630.23	0
630	Ochratoxin A (ppb)	0563	60.94	4.320	55.58	18.35	4.470	44	0.29	630.20	0
630	Ochratoxin A (ppb)	2352	61.59	0.8100	55.58	18.35	4.470	44	0.33	630.23	0
630	Ochratoxin A (ppb)	2140	62.00	10.00	55.58	18.35	4.470	44	0.35	630.23	0
630	Ochratoxin A (ppb)	2322	62.13	0.0000	55.58	18.35	4.470	44	0.36	630.20	0
630	Ochratoxin A (ppb)	2192	64.71	6.350	55.58	18.35	4.470	44	0.50	630.20	0
630	Ochratoxin A (ppb)	0218	65.85	4.900	55.58	18.35	4.470	44	0.56	630.23	0
630	Ochratoxin A (ppb)	0297	68.00	4.000	55.58	18.35	4.470	44	0.68	630.23	0
630	Ochratoxin A (ppb)	2022	76.79	12.06	55.58	18.35	4.470	44	1.16	630.03	0
630	Ochratoxin A (ppb)	0148	77.21	19.13	55.58	18.35	4.470	44	1.18	630.23	0
630	Ochratoxin A (ppb)	0913	77.40	1.600	55.58	18.35	4.470	44	1.19	630.23	0
630	Ochratoxin A (ppb)	2055	80.65	4.700	55.58	18.35	4.470	44	1.37	630.23	0
630	Ochratoxin A (ppb)	0098	83.90	22.60	55.58	18.35	4.470	44	1.54	630.20	0
630	Ochratoxin A (ppb)	2129	88.45	1.100	55.58	18.35	4.470	44	1.79	630.23	0
630	Ochratoxin A (ppb)	2327	91.10	0.6000	55.58	18.35	4.470	44	1.94	630.23	0
630	Ochratoxin A (ppb)	0674	96.10	0.0000	55.58	18.35	4.470	44	2.21	630.20	0
630	Ochratoxin A (ppb)	0529	292.4	3.300	55.58	18.35	4.470	44	12.90	630.01	2
640	T-2 Toxin (ppb)	0656	48.56	0.0000	104.1	33.79	10.33	23	-1.64	640.05	0
640	T-2 Toxin (ppb)	0964	54.50	5.000	104.1	33.79	10.33	23	-1.47	640.23	0
640	T-2 Toxin (ppb)	2326	63.50	25.00	104.1	33.79	10.33	23	-1.20	640.23	0
640	T-2 Toxin (ppb)	0910	68.06	0.0000	104.1	33.79	10.33	23	-1.07	640.21	0
640	T-2 Toxin (ppb)	2033	71.00	6.000	104.1	33.79	10.33	23	-0.98	640.23	0
640	T-2 Toxin (ppb)	2140	78.50	3.000	104.1	33.79	10.33	23	-0.76	640.23	0
640	T-2 Toxin (ppb)	0913	88.40	4.400	104.1	33.79	10.33	23	-0.46	640.23	0
640	T-2 Toxin (ppb)	0202	91.70	14.80	104.1	33.79	10.33	23	-0.37	640.23	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests			
640	T-2 Toxin (ppb)	2239	92.25	1.700	104.1	33.79	10.33	23	-0.35	640.98	0
640	T-2 Toxin (ppb)	2302	99.42	2.710	104.1	33.79	10.33	23	-0.14	640.99	0
640	T-2 Toxin (ppb)	0870	103.9	4.220	104.1	33.79	10.33	23	-0.01	640.23	0
640	T-2 Toxin (ppb)	0682	107.2	0.0000	104.1	33.79	10.33	23	0.09	640.23	0
640	T-2 Toxin (ppb)	2369	107.5	2.640	104.1	33.79	10.33	23	0.10	640.23	0
640	T-2 Toxin (ppb)	0353	115.5	3.000	104.1	33.79	10.33	23	0.34	640.23	0
640	T-2 Toxin (ppb)	0407	116.0	42.40	104.1	33.79	10.33	23	0.35	640.05	0
640	T-2 Toxin (ppb)	0563	120.0	14.00	104.1	33.79	10.33	23	0.47	640.25	0
640	T-2 Toxin (ppb)	0218	120.5	15.00	104.1	33.79	10.33	23	0.49	640.23	0
640	T-2 Toxin (ppb)	0148	123.5	14.90	104.1	33.79	10.33	23	0.57	640.23	0
640	T-2 Toxin (ppb)	2129	135.0	6.000	104.1	33.79	10.33	23	0.92	640.23	0
640	T-2 Toxin (ppb)	0674	140.7	0.0000	104.1	33.79	10.33	23	1.08	640.03	0
640	T-2 Toxin (ppb)	2327	143.8	8.500	104.1	33.79	10.33	23	1.17	640.23	0
640	T-2 Toxin (ppb)	2352	144.8	18.82	104.1	33.79	10.33	23	1.20	640.23	0
640	T-2 Toxin (ppb)	0066	157.5	34.30	104.1	33.79	10.33	23	1.58	640.03	0
641	HT-2 Toxin (ppb)	0353	10.00	2.000	18.35	6.244	2.401	9	-1.34	641.23	0
641	HT-2 Toxin (ppb)	2033	11.50	1.000	18.35	6.244	2.401	9	-1.10	641.23	0
641	HT-2 Toxin (ppb)	2369	13.01	0.4000	18.35	6.244	2.401	9	-0.86	641.23	0
641	HT-2 Toxin (ppb)	0910	13.15	0.0000	18.35	6.244	2.401	9	-0.83	641.21	0
641	HT-2 Toxin (ppb)	2326	13.50	1.000	18.35	6.244	2.401	9	-0.78	641.23	0
641	HT-2 Toxin (ppb)	2327	19.25	4.100	18.35	6.244	2.401	9	0.14	641.23	0
641	HT-2 Toxin (ppb)	0148	21.17	1.640	18.35	6.244	2.401	9	0.45	641.23	0
641	HT-2 Toxin (ppb)	2352	30.90	3.630	18.35	6.244	2.401	9	2.01	641.23	0
641	HT-2 Toxin (ppb)	0202	50.30	16.00	18.35	6.244	2.401	9	5.12	641.23	0
641	HT-2 Toxin (ppb)	0913	< 5		18.35	6.244	2.401	9		641.23	5
641	HT-2 Toxin (ppb)	0218	< 20		18.35	6.244	2.401	9		641.23	5
641	HT-2 Toxin (ppb)	0870	< 25		18.35	6.244	2.401	9		641.23	5
641	HT-2 Toxin (ppb)	0964	< 25		18.35	6.244	2.401	9		641.23	5
641	HT-2 Toxin (ppb)	2140	< 200		18.35	6.244	2.401	9		641.23	5
642	T-2 Toxin + HT-2 Toxin (ppb)	0596	45.90	27.40	134.1	43.25	10.31	15	-2.04	642.98	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0505	99.92	18.69	134.1	43.25	10.31	15	-0.79	642.16	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0590	107.5	5.000	134.1	43.25	10.31	15	-0.62	642.98	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0407	116.6	2.430	134.1	43.25	10.31	15	-0.40	642.99	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2369	120.5	3.040	134.1	43.25	10.31	15	-0.32	642.04	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2022	120.5	6.490	134.1	43.25	10.31	15	-0.32	642.98	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0629	122.5	25.00	134.1	43.25	10.31	15	-0.27	642.98	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0353	123.0	10.00	134.1	43.25	10.31	15	-0.26	642.04	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2022	129.8	8.400	134.1	43.25	10.31	15	-0.10	642.16	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0202	142.0	2.000	134.1	43.25	10.31	15	0.18	642.04	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2022	157.2	3.620	134.1	43.25	10.31	15	0.53	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	2022	165.2	0.0000	134.1	43.25	10.31	15	0.72	642.99	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0227	169.5	3.000	134.1	43.25	10.31	15	0.82	642.21	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests			
642	T-2 Toxin + HT-2 Toxin (ppb)	2192	169.5	9.000	134.1	43.25	10.31	15	0.82	642.21	0
642	T-2 Toxin + HT-2 Toxin (ppb)	0529	275.8	22.50	134.1	43.25	10.31	15	3.27	642.21	0
650	Zearalenone (ppb)	0520	27.50	1.000	132.0	42.57	10.55	49	-2.45	650.99	0
650	Zearalenone (ppb)	0066	56.90	2.400	132.0	42.57	10.55	49	-1.76	650.07	0
650	Zearalenone (ppb)	0964	60.00	6.000	132.0	42.57	10.55	49	-1.69	650.24	0
650	Zearalenone (ppb)	0553	73.55	5.300	132.0	42.57	10.55	49	-1.37	650.24	0
650	Zearalenone (ppb)	2033	92.50	15.00	132.0	42.57	10.55	49	-0.93	650.24	0
650	Zearalenone (ppb)	0590	93.85	14.30	132.0	42.57	10.55	49	-0.90	650.06	0
650	Zearalenone (ppb)	2326	98.00	6.000	132.0	42.57	10.55	49	-0.80	650.24	0
650	Zearalenone (ppb)	2302	99.25	1.150	132.0	42.57	10.55	49	-0.77	650.99	0
650	Zearalenone (ppb)	2369	101.0	1.460	132.0	42.57	10.55	49	-0.73	650.20	0
650	Zearalenone (ppb)	2369	101.0	1.460	132.0	42.57	10.55	49	-0.73	650.21	0
650	Zearalenone (ppb)	0354	103.3	1.970	132.0	42.57	10.55	49	-0.67	650.21	0
650	Zearalenone (ppb)	0629	106.0	6.000	132.0	42.57	10.55	49	-0.61	650.20	0
650	Zearalenone (ppb)	0674	113.0	0.0000	132.0	42.57	10.55	49	-0.45	650.20	0
650	Zearalenone (ppb)	0913	113.2	17.80	132.0	42.57	10.55	49	-0.44	650.24	0
650	Zearalenone (ppb)	2140	114.0	38.00	132.0	42.57	10.55	49	-0.42	650.24	0
650	Zearalenone (ppb)	0870	114.1	1.671	132.0	42.57	10.55	49	-0.42	650.01	0
650	Zearalenone (ppb)	2129	117.5	9.000	132.0	42.57	10.55	49	-0.34	650.24	0
650	Zearalenone (ppb)	0563	118.9	7.030	132.0	42.57	10.55	49	-0.31	650.20	0
650	Zearalenone (ppb)	2327	119.7	26.70	132.0	42.57	10.55	49	-0.29	650.24	0
650	Zearalenone (ppb)	0910	120.0	8.000	132.0	42.57	10.55	49	-0.28	650.20	0
650	Zearalenone (ppb)	2196	121.4	0.0000	132.0	42.57	10.55	49	-0.25	650.20	0
650	Zearalenone (ppb)	0218	131.0	0.2000	132.0	42.57	10.55	49	-0.02	650.24	0
650	Zearalenone (ppb)	0656	131.8	0.0000	132.0	42.57	10.55	49	0.00	650.06	0
650	Zearalenone (ppb)	2022	132.0	2.160	132.0	42.57	10.55	49	0.00	650.06	0
650	Zearalenone (ppb)	2352	133.8	34.84	132.0	42.57	10.55	49	0.04	650.24	0
650	Zearalenone (ppb)	2066	134.0	4.000	132.0	42.57	10.55	49	0.05	650.06	0
650	Zearalenone (ppb)	0505	135.2	4.770	132.0	42.57	10.55	49	0.08	650.16	0
650	Zearalenone (ppb)	0035	136.6	2.300	132.0	42.57	10.55	49	0.11	650.06	0
650	Zearalenone (ppb)	2192	142.5	5.000	132.0	42.57	10.55	49	0.25	650.20	0
650	Zearalenone (ppb)	2226	144.2	10.23	132.0	42.57	10.55	49	0.29	650.06	0
650	Zearalenone (ppb)	0148	146.9	4.620	132.0	42.57	10.55	49	0.35	650.24	0
650	Zearalenone (ppb)	0407	147.3	47.15	132.0	42.57	10.55	49	0.36	650.07	0
650	Zearalenone (ppb)	2022	147.6	1.080	132.0	42.57	10.55	49	0.37	650.05	0
650	Zearalenone (ppb)	0596	149.3	8.300	132.0	42.57	10.55	49	0.41	650.98	0
650	Zearalenone (ppb)	2196	150.0	0.0000	132.0	42.57	10.55	49	0.42	650.04	0
650	Zearalenone (ppb)	0098	150.0	0.0000	132.0	42.57	10.55	49	0.42	650.20	0
650	Zearalenone (ppb)	2022	152.4	22.85	132.0	42.57	10.55	49	0.48	650.04	0
650	Zearalenone (ppb)	2022	155.8	7.620	132.0	42.57	10.55	49	0.56	650.01	0
650	Zearalenone (ppb)	0529	156.7	8.400	132.0	42.57	10.55	49	0.58	650.01	0
650	Zearalenone (ppb)	0202	158.5	5.000	132.0	42.57	10.55	49	0.62	650.24	0

Method Group	Analyte Group (Units)	Lab Code	Lab Data		Robust* Analyte Values				AAFCO PT Z Score	Your Method	Flag
			Value	Range	Rob. Mean	Rev.Horwitz	Rob. R-bar	# Tests			
650	Zearalenone (ppb)	0297	160.0	2.000	132.0	42.57	10.55	49	0.66	650.24	0
650	Zearalenone (ppb)	0010	162.0	24.00	132.0	42.57	10.55	49	0.71	650.24	0
650	Zearalenone (ppb)	0009	164.1	23.68	132.0	42.57	10.55	49	0.76	650.24	0
650	Zearalenone (ppb)	0027	173.3	91.60	132.0	42.57	10.55	49	0.97	650.01	0
650	Zearalenone (ppb)	0353	173.5	61.00	132.0	42.57	10.55	49	0.98	650.23	0
650	Zearalenone (ppb)	2239	178.0	4.000	132.0	42.57	10.55	49	1.08	650.20	0
650	Zearalenone (ppb)	0354	180.4	42.40	132.0	42.57	10.55	49	1.14	650.04	0
650	Zearalenone (ppb)	2239	183.5	13.00	132.0	42.57	10.55	49	1.21	650.98	0
650	Zearalenone (ppb)	2268	220.0	20.00	132.0	42.57	10.55	49	2.07	650.99	0
650	Zearalenone (ppb)	0242	1,279	127.0	132.0	42.57	10.55	49	26.93	650.24	2
650	Zearalenone (ppb)	0682	< 100		132.0	42.57	10.55	49		650.24	5

Note 1: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or between -2 and -3 (warning) and Green = Z <= 2 and >= -2 (OK at 95%). Horwitz SD's assigned based on Rob Mean only.

Note 2: Data Not Used Flags: 1=Rejected for duplicates too far apart, 2=Rejected as extreme outlier, 5=Reporting limit (<), 4=Zeros submitted as values and 3=Statistical problem. Flag 0 indicates data used in calculations.

Note 3: *Robust statistics not used if < 6 labs reporting, in the case of 4 or 5 labs reporting Means and SD's may be reported based on Raw Data with obvious blunders removed (Mandel h and k exclusions apply; Grey). A zero range is not included in robust R-bar calculation.