

Method

Summary Statistics



Statistical reports have been updated to a new organizational format.

Statistical analysis and number of reports remain the same. The types of reports are shown below.

Results from all labs...

- ...sorted by analyte (Analyte All Labs PT Report)

- ...sorted by method (Method All Labs PT Report)

Summary statistics...

- ...for each analyte (Analyte Summary Statistics)

- ...for each method (**Method Summary Statistics**)

Report cards evaluating individual lab performance...

- ...for an analyte regardless of method (Analyte Laboratory Report Card)

- ...for a specific method (Method Laboratory Report Card).

Detailed description on the content of the **Method Summary Statistics** is provided in the Appendix



METHOD Summary Statistics

202623 (Chick Starter medicated)

Issue Date: 4/30/2026

Code	Analyte / Method	¹ Trueness (Lab Value)				² Thompson Horwitz %RSD	³ Precision (range)		
		Robust Mean	n used	Robust Uncert.	Robust StDev		Robust Mean	n used	
001.07	Loss on Drying/104°C 3 hr, in malt (%)	11.1	34	0.0985	0.4594	4.14	2.78	0.1444	31
001.99	Loss on Drying/Miscellaneous (%)	10.79	12	0.1158	0.3209	2.97	2.79	0.0846	12
001.00	Loss on Drying/Vac 95°C 5 hr (%)	10.97	4	0.5502	0.8804	8.03	2.79	0.0977	4
001.03	Loss on Drying/Low temp. methods (%)	11.6	1					0.2	1
002.06	Protein, Crude/Combustion Nitrogen Analyzer (%)	18.73	91	0.0427	0.326	1.74	2.31	0.2588	85
002.01	Protein, Crude/Auto Kjel-Foss (%)	18.6	12	0.0998	0.2766	1.49	2.32	0.166	9
002.05	Protein, Crude/Copper, Boric Acid (%)	18.52	8	0.1291	0.2922	1.58	2.32	0.1994	8
002.11	Protein, Crude/NIR (%)	17.22	3	1.699	2.354	13.7	2.41	0.0833	3
002.02	Protein, Crude/Semiauto Autoanalyzer (%)	18.76	2		0.2652			0.135	2
002.04	Protein, Crude/Copper Catalyst (%)	18.2	2		0.8415			0.7	2
002.99	Protein, Crude/Miscellaneous (%)	18.74	2		0.5515			0.07	2
002.08	Protein, Crude/Cu/Ti (%)	18.53	2		0.0681			0.11785	2
002.00	Protein, Crude/Crude (%)	17.33	1					0.24	1

Code	Analyte / Method	¹ Trueness (Lab Value)				² Thompson Horwitz %RSD	³ Precision (range)		
		Robust Mean	n used	Robust Uncert.	Robust StDev		Robust Mean	n used	
003.14	Fat, Crude/Ankom (%)	2.405	45	0.0424	0.2275	9.46	3.5	0.1152	43
003.10	Fat, Crude/Randall, Pet Ether (%)	2.388	24	0.0421	0.1652	6.92	3.5	0.0996	22
003.06	Fat, Crude/Pet Ether (%)	2.316	10	0.0827	0.2093	9.04	3.52	0.0936	9
003.09	Fat, Crude/Randall, Diethyl Ether Ext (%)	2.401	7	0.0351	0.0743	3.09	3.5	0.0906	7
003.99	Fat, Crude/Miscellaneous (%)	2.171	4	0.2465	0.3944	18.2	3.55	0.1725	4
003.00	Fat, Crude/Diethyl Ether Ext., Direct (%)	2.542	4	0.1965	0.3144	12.4	3.47	0.115	4
003.13	Fat, Crude/Randall, Hexane Ext. (%)	2.404	3	0.1021	0.1415	5.89	3.5	0.2886	3
003.01	Fat, Crude/Diethyl Ether Ext (13th ed.), Indirect (%)	2.471	3	0.1127	0.1561	6.32	3.48	0.1514	3
003.12	Fat, Crude/Hexane Ext (%)	2.395	3	0.1528	0.2117	8.84	3.5	0.03	3
003.11	Fat, Crude/NIR (%)	2.622	2		0.9864			0.035	2
004.07	Fiber, Crude/ANKOM (%)	3.628	68	0.0528	0.3481	9.59	3.29	0.1497	63
004.06	Fiber, Crude/Fibertec (%)	3.95	11	0.1248	0.3312	8.39	3.25	0.1395	11
004.00	Fiber, Crude/Asbestos Free (%)	3.669	8	0.1506	0.3407	9.29	3.28	0.0888	8
004.99	Fiber, Crude/Miscellaneous (%)	3.748	3	0.4847	0.6716	17.9	3.27	0.1123	3
004.11	Fiber, Crude/NIR (%)	4.242	2		0.145			0.075	2
004.03	Fiber, Crude/Fritted Glass (%)	3.502	2		0.6128			0.3201	2
005.00	Ash/2h @ 600°C (%)	5.679	80	0.037	0.265	4.67	3.08	0.1089	76
005.05	Ash/3h @ 550°C (%)	5.887	16	0.0672	0.215	3.65	3.06	0.0652	16
005.99	Ash/Miscellaneous (%)	5.592	5	0.2557	0.4574	8.18	3.08	0.084	5

Code	Analyte / Method	¹ Trueness (Lab Value)					² Thompson Horwitz %RSD	³ Precision (range)	
		Robust Mean	n used	Robust Uncert.	Robust StDev	Robust %RSD		Robust Mean	n used
005.11	Ash/NIR (%)	8.067	3	1.253	1.736	21.5	2.92		
005.03	Ash/Microwave furnace (%)	5.626	3	0.1985	0.275	4.89	3.08	0.0513	3
006.99	Total Sugars/Miscellaneous (%)	4.179	5	0.7804	1.396	33.4	3.22	0.142	5
006.00	Total Sugars/As sucrose (%)	7.853	2		6.92			0.27015	2
006.01	Total Sugars/Mod. Fehling Soln (%)	5.225	1					0.15	1
008.08	Fiber, Acid Detergent/Filter Bag - ANKOM (%)	5.389	37	0.1306	0.6356	11.8	3.1	0.1872	32
008.02	Fiber, Acid Detergent/Crucible (%)	5.813	7	0.3955	0.8371	14.4	3.06	0.1365	7
008.99	Fiber, Acid Detergent/Miscellaneous (%)	6.475	1					0.37	1
009.09	Fiber, Neutral Detergent/Filter Bag - ANKOM (%)	13.91	35	0.2228	1.054	7.58	2.68	0.2937	31
009.07	Fiber, Neutral Detergent/AOAC -ENZ Pretreat (%)	15.09	6	0.5004	0.9806	6.5	2.57	0.3938	4
010.99	Moisture/Miscellaneous (18%)	11.2	16	0.2108	0.6745	6.02	2.78	0.193	16
010.11	Moisture/NIR (18%)	11.26	2		0.4773			0.125	2
010.03	Moisture/Karl-Fischer (18%)	11.42	1					0.44	1
011.01	Loss on Drying/135°C 2hr (%)	12.04	50	0.066	0.3732	3.1	2.75	0.1049	47
011.02	Loss on Drying/130°C for 2 hours (%)	11.9	2		0.0212			0.07	2
011.99	Loss on Drying/High Temp. Methods Miscellaneous (%)	10.85	1					0.4	1
012.04	Starch/Enzymatic-Enzyme Membrane Technology (YSI) (%)	36.68	8	0.7597	1.719	4.69	1.65	0.7012	8
012.01	Starch/Enzymatic-Colorimetric Method (Megazyme) (%)	36.17	7	1.683	3.563	9.85	1.66	0.9105	7
012.00	Starch/Polarimetric (Ewers) (%)	38.73	6	0.271	0.5311	1.37	1.61	0.1167	3

Code	Analyte / Method	¹ Trueness (Lab Value)				² Thompson Horwitz %RSD	³ Precision (range)		
		Robust Mean	n used	Robust Uncert.	Robust StDev		Robust Mean	n used	
012.11	Starch/NIR (%)	36.81	2		1.008		0.255	2	
012.99	Starch/Miscellaneous (%)	32.14	2		4.922		0.39525	2	
012.03	Starch/Enzymatic-Colorimetric Method, Miscellaneous (%)	37.59	2		1.804		0.6043	2	
012.20	Starch/Dietary, Enzymatic-Colorimetric (%)	34.96	1				0.52	1	
013.00	Fat, Pretreat/Acid hydrolysis (%)	3.217	14	0.1274	0.3812	11.9	3.35	0.132	13
013.02	Fat, Pretreat/Mojonnier, Bak Ext (%)	3.736	11	0.2381	0.6316	16.9	3.27	0.265	11
013.13	Fat, Pretreat/Ankom- Acid Hydrolysis (%)	2.776	8	0.1467	0.3318	12	3.42	0.116	8
013.10	Fat, Pretreat/Soxtec-Acid Hydrolysis (%)	2.819	5	0.3035	0.5429	19.3	3.42	0.3967	3
013.08	Fat, Pretreat/Roese-Gottlieb Modified (%)	1.485	1					0.05	1
014.02	Fiber, Total Dietary/ANKOM Enz-Grav (%)	17.25	1					0.3	1
015.43	Aluminum/ICP, Microwave (ppm)	84.26	6	7.686	15.06	17.9	8.17	1.293	6
015.41	Aluminum/ICP, Dry ash (ppm)	73.71	3	7.952	11.02	14.9	8.33	3.414	3
015.42	Aluminum/ICP, Open vessel (ppm)	34.78	2		10.3			1.51	2
015.53	Aluminum/ICP-MS, Microwave (ppm)	83.14	2		0.2739			7.51995	2
015.99	Aluminum/Miscellaneous (ppm)	46.75	1					0.7	1
017.43	Boron/ICP, Microwave (ppm)	8.21	6	0.4962	0.9724	11.8	11.6	0.1431	4
017.42	Boron/ICP, Open vessel (ppm)	9.159	4	1.383	2.213	24.2	11.4	0.729	4
017.41	Boron/ICP, Dry ash (ppm)	8.48	4	0.3309	0.5294	6.24	11.5	0.5662	4
017.53	Boron/ICP-MS, Microwave (ppm)	9.15	1					1.3134	1

Code	Analyte / Method	¹ Trueness (Lab Value)					² Thompson Horwitz %RSD	³ Precision (range)	
		Robust Mean	n used	Robust Uncert.	Robust StDev	Robust %RSD		Robust Mean	n used
017.99	Boron/Miscellaneous (ppm)	8.69	1					0.02	1
019.43	Calcium/ICP, Microwave (%)	1.005	34	0.0114	0.0531	5.28	3.99	0.0376	32
019.42	Calcium/ICP, Open vessel (%)	0.9933	18	0.0157	0.0534	5.37	3.99	0.0304	16
019.41	Calcium/ICP, Dry ash (%)	0.979	17	0.0114	0.0375	3.83	4	0.0384	17
019.31	Calcium/AAS, Dry ash (%)	1.045	12	0.0319	0.0884	8.46	3.96	0.0175	10
019.99	Calcium/Miscellaneous (%)	0.8702	6	0.0773	0.1514	17.4	4.07	0.0416	6
019.08	Calcium/EDTA (%)	1.068	6	0.0192	0.0376	3.52	3.95	0.0267	3
019.00	Calcium/Ox-Mn04 Vol. (%)	1.005	5	0.1121	0.2006	20	3.99	0.0182	5
019.52	Calcium/ICP-MS, Open vessel (%)	0.9964	3	0.1123	0.1557	15.6	3.99	0.0731	3
019.53	Calcium/ICP-MS, Microwave (%)	1.078	2		0.1374			0.2182	2
019.44	Calcium/ICP, Dry ash (%)	0.9488	2		0.0477			0.0315	2
019.09	Calcium/Ion-selective electrode (%)	0.9595	1					0.021	1
019.32	Calcium/AAS, Open vessel (%)	1.015	1					0.05	1
021.43	Cobalt/ICP, Microwave (ppm)	0.5657	6	0.1253	0.2456	43.4	17.3	0.0766	5
021.53	Cobalt/ICP-MS, Microwave (ppm)	0.5331	5	0.0188	0.0336	6.31	17.5	0.0347	5
021.42	Cobalt/ICP, Open vessel (ppm)	0.3765	2		0.1252			0.068	2
021.52	Cobalt/ICP-MS, Open vessel (ppm)	0.4856	1					0.0253	1
021.31	Cobalt/AAS, Dry ash (ppm)	0.4855	1					0.002	1
021.41	Cobalt/ICP, Dry ash (ppm)	0.8046	1					0.0118	1

Code	Analyte / Method	¹ Trueness (Lab Value)					² Thompson Horwitz %RSD	³ Precision (range)	
		Robust Mean	n used	Robust Uncert.	Robust StDev	Robust %RSD		Robust Mean	n used
022.43	Copper/ICP, Microwave (ppm)	9.883	29	0.2201	0.9481	9.59	11.3	0.6551	27
022.42	Copper/ICP, Open vessel (ppm)	10.1	16	0.2948	0.9435	9.34	11.2	0.3187	15
022.41	Copper/ICP, Dry ash (ppm)	10.45	10	0.4967	1.257	12	11.2	0.5716	10
022.31	Copper/AAS, Dry ash (ppm)	10.56	4	0.8404	1.345	12.7	11.2	0.057	3
022.44	Copper/ICP, Dry ash (ppm)	9.357	3	0.8455	1.172	12.5	11.4	0.38	3
022.99	Copper/Miscellaneous (ppm)	11.43	3	2.356	3.265	28.6	11		
022.52	Copper/ICP-MS, Open vessel (ppm)	10.41	2		0.5904			0.235	2
022.53	Copper/ICP-MS, Microwave (ppm)	9.799	2		0.6453			1.55425	2
022.32	Copper/AAS, Open vessel (ppm)	11.9	1					1.2	1
024.99	Iodine/Miscellaneous (ppm)	0.6085	1					0.033	1
025.43	Iron/ICP, Microwave (ppm)	136.1	28	2.155	9.124	6.7	7.6	3.967	27
025.42	Iron/ICP, Open vessel (ppm)	126.2	16	3.178	10.17	8.06	7.69	5.117	14
025.41	Iron/ICP, Dry ash (ppm)	127.8	14	3.203	9.586	7.5	7.67	2.861	14
025.31	Iron/AAS, Dry ash (ppm)	142.3	7	5.359	11.34	7.97	7.55	5.963	6
025.99	Iron/Miscellaneous (ppm)	124.9	4	12.7	20.32	16.3	7.7	4.527	3
025.52	Iron/ICP-MS, Open vessel (ppm)	116.1	2		3.507			8.57	2
025.53	Iron/ICP-MS, Microwave (ppm)	124.9	2		9.347			#####	2
027.43	Magnesium/ICP, Microwave (%)	0.2123	30	0.0026	0.0112	5.27	5.03	0.0061	23
027.42	Magnesium/ICP, Open vessel (%)	0.2142	18	0.0035	0.012	5.62	5.03	0.0054	12

Code	Analyte / Method	¹ Trueness (Lab Value)				² Thompson Horwitz %RSD	³ Precision (range)		
		Robust Mean	n used	Robust Uncert.	Robust StDev		Robust Mean	n used	
027.41	Magnesium/ICP, Dry ash (%)	0.2068	13	0.0032	0.0092	4.46	5.05	0.0049	11
027.31	Magnesium/AAS, Dry ash (%)	0.2125	7	0.0064	0.0135	6.36	5.03	0.0052	5
027.99	Magnesium/Miscellaneous (%)	0.2567	5	0.0512	0.0916	35.7	4.89	0.0152	4
027.52	Magnesium/ICP-MS, Open vessel (%)	0.2237	3	0.0138	0.0191	8.54	5	0.0126	3
027.44	Magnesium/ICP, Dry ash (%)	0.2028	2		0.011			0.0105	2
027.53	Magnesium/ICP-MS, Microwave (%)	0.2106	2		0.0008			0.02	2
027.32	Magnesium/AAS, Open vessel (%)	0.215	1					0.01	1
028.43	Manganese/ICP, Microwave (ppm)	87.81	29	1.175	5.06	5.76	8.12	3.645	27
028.42	Manganese/ICP, Open vessel (ppm)	86.48	18	1.927	6.54	7.56	8.14	4.821	18
028.41	Manganese/ICP, Dry ash (ppm)	82.33	10	1.494	3.781	4.59	8.2	2.298	10
028.31	Manganese/AAS, Dry ash (ppm)	86.4	8	2.035	4.605	5.33	8.14	1.852	7
028.99	Manganese/Miscellaneous (ppm)	88	4	4.059	6.494	7.38	8.12	3.19	4
028.44	Manganese/ICP, Dry ash (ppm)	80.6	3	6.809	9.435	11.7	8.22	0.2433	3
028.53	Manganese/ICP-MS, Microwave (ppm)	81.46	3	7.289	10.1	12.4	8.21	6.152	3
028.52	Manganese/ICP-MS, Open vessel (ppm)	84.32	2		1.75			6.805	2
028.32	Manganese/AAS, Open vessel (ppm)	92.95	1					1.3	1
031.43	Phosphorus/ICP, Microwave (%)	0.6406	33	0.0052	0.024	3.75	4.27	0.0182	30
031.42	Phosphorus/ICP, Open vessel (%)	0.6305	19	0.0106	0.037	5.87	4.28	0.018	16
031.01	Phosphorus/Photometric (%)	0.6338	18	0.0077	0.0261	4.12	4.27	0.0104	10

Code	Analyte / Method	¹ Trueness (Lab Value)				² Thompson Horwitz %RSD	³ Precision (range)		
		Robust Mean	n used	Robust Uncert.	Robust StDev		Robust %RSD	Robust Mean	n used
031.41	Phosphorus/ICP, Dry ash (%)	0.6192	16	0.0067	0.0214	3.46	4.29	0.0187	15
031.99	Phosphorus/Miscellaneous (%)	0.6376	6	0.0224	0.044	6.9	4.27	0.0182	6
031.44	Phosphorus/ICP, Dry ash (%)	0.6173	3	0.0183	0.0253	4.1	4.29		
031.52	Phosphorus/ICP-MS, Open vessel (%)	0.6511	3	0.0262	0.0363	5.57	4.26	0.0301	3
031.53	Phosphorus/ICP-MS, Microwave (%)	0.6268	2		0.0103			0.0464	2
031.03	Phosphorus/Autoanalyzer (%)	0.6388	2		0.0088			0.0305	2
031.02	Phosphorus/GQMP (AOAC 935.13-Extraction) (%)	0.6157	1					0.0488	1
032.43	Potassium/ICP, Microwave (%)	0.9803	31	0.0109	0.0486	4.96	4	0.0215	30
032.42	Potassium/ICP, Open vessel (%)	0.9979	17	0.0183	0.0603	6.05	3.99	0.0227	16
032.41	Potassium/ICP, Dry ash (%)	0.9405	14	0.0165	0.0493	5.24	4.03	0.0291	14
032.31	Potassium/AAS, Dry ash (%)	0.9455	6	0.0339	0.0665	7.03	4.02	0.0245	5
032.99	Potassium/Miscellaneous (%)	1.001	5	0.0505	0.0903	9.02	3.99	0.035	4
032.52	Potassium/ICP-MS, Open vessel (%)	0.9788	3	0.0826	0.1144	11.7	4	0.0522	3
032.44	Potassium/ICP, Dry ash (%)	0.946	2		0.0721			0.038	2
032.32	Potassium/AAS, Open vessel (%)	1.075	1					0.01	1
032.53	Potassium/ICP-MS, Microwave (%)	0.9754	1					0.1548	1
033.01	Salt as chloride/Poten Cl (%)	0.4561	21	0.0046	0.0169	3.71	4.49	0.0166	15
033.99	Salt/Miscellaneous (%)	0.4468	13	0.013	0.0376	8.42	4.5	0.014	12
033.00	Salt as chloride/Sol Cl (%)	0.4743	13	0.014	0.0404	8.52	4.46	0.019	9

Code	Analyte / Method	¹ Trueness (Lab Value)				² Thompson Horwitz %RSD	³ Precision (range)	
		Robust Mean	n used	Robust Uncert.	Robust StDev		Robust Mean	n used
033.03	Salt as chloride/Quantab (%)	0.44	4	0.0319	0.051	11.6	4.51	
033.05	Salt as chloride/Ion Sel Electrode (%)	0.3933	3	0.0235	0.0325	8.27	4.59	
034.53	Selenium/ICP-MS, Microwave (ppm)	0.295	7	0.0225	0.0476	16.1	19.1	0.0425 7
034.04	Selenium/AA, Hydride (ppm)	0.2643	3	0.0323	0.0448	17	19.4	0.0146 3
034.52	Selenium/ICP-MS, Open vessel (ppm)	0.3049	2		0.0211			0.01715 2
034.43	Selenium/ICP, Microwave (ppm)	0.279	2		0.0509			0.014 2
034.42	Selenium/ICP, Open vessel (ppm)	1.025	1					0.73 1
034.01	Selenium/Fluor (ppm)	0.28	1					
034.99	Selenium/Miscellaneous (ppm)	0.6255	1					0.001 1
035.43	Sodium/ICP, Microwave (%)	0.1648	29	0.0023	0.0098	5.92	5.23	0.0063 24
035.41	Sodium/ICP, Dry ash (%)	0.1707	18	0.0036	0.0122	7.15	5.2	0.0081 15
035.42	Sodium/ICP, Open vessel (%)	0.1662	16	0.0033	0.0106	6.37	5.22	0.004 10
035.31	Sodium/AAS, Dry ash (%)	0.183	5	0.008	0.0143	7.81	5.15	0.0049 3
035.99	Sodium/Miscellaneous (%)	0.4027	5	0.1969	0.3523	87.5	4.57	0.0222 4
035.52	Sodium/ICP-MS, Open vessel (%)	0.1649	3	0.0127	0.0177	10.7	5.23	0.0078 3
035.53	Sodium/ICP-MS, Microwave (%)	0.1696	2		0.0079			0.00755 2
035.32	Sodium/AAS, Open vessel (%)	0.19	1					0.04 1
035.05	Sodium/Flame Emission (%)	0.155	1					0.01 1
036.43	Sulfur/ICP, Microwave (%)	0.2481	26	0.0038	0.0154	6.22	4.92	0.0066 20

Code	Analyte / Method	¹ Trueness (Lab Value)				² Thompson Horwitz %RSD	³ Precision (range)		
		Robust Mean	n used	Robust Uncert.	Robust StDev		Robust Mean	n used	
036.42	Sulfur/ICP, Open vessel (%)	0.2398	18	0.0052	0.0177	7.4	4.94	0.0088	12
036.04	Sulfur/LECO (%)	0.2405	4	0.0117	0.0187	7.79	4.94		
036.52	Sulfur/ICP-MS, Open vessel (%)	0.2852	2		0.057			0.05605	2
036.99	Sulfur/Miscellaneous (%)	0.2175	2		0.0177			0.01	1
036.00	Sulfur/Gravimetric (%)	0.2385	1					0.033	1
036.53	Sulfur/ICP-MS, Microwave (%)	0.1745	1					0.005	1
037.43	Zinc/ICP, Microwave (ppm)	83.08	32	1.753	7.933	9.55	8.19	3.289	31
037.42	Zinc/ICP, Open vessel (ppm)	81.26	17	2.89	9.534	11.7	8.21	4.295	17
037.41	Zinc/ICP, Dry ash (ppm)	82.37	12	2.905	8.05	9.77	8.2	3.74	12
037.31	Zinc/AAS, Dry ash (ppm)	83.73	5	3.278	5.863	7	8.18	1.265	3
037.99	Zinc/Miscellaneous (ppm)	78.15	4	3.971	6.353	8.13	8.26	3.388	4
037.44	Zinc/ICP, Dry ash (ppm)	72.92	3	5.145	7.129	9.78	8.35	2.647	3
037.53	Zinc/ICP-MS, Microwave (ppm)	84.26	3	5.4	7.482	8.88	8.17	5.906	3
037.52	Zinc/ICP-MS, Open vessel (ppm)	84.72	2		1.011			6.12	2
037.32	Zinc/AAS, Open vessel (ppm)	87.8	1					2	1
038.43	Molybdenum/ICP, Microwave (ppm)	0.7923	7	0.1583	0.3351	42.3	16.4	0.0997	7
038.42	Molybdenum/ICP, Open vessel (ppm)	1.001	6	0.1024	0.2007	20	15.9	0.2203	6
038.53	Molybdenum/ICP-MS, Microwave (ppm)	1.022	6	0.0648	0.127	12.4	15.8	0.0625	5
038.41	Molybdenum/ICP, Dry ash (ppm)	1.004	2		0.1922			0.03395	2

Code	Analyte / Method	¹ Trueness (Lab Value)				² Thompson Horwitz %RSD	³ Precision (range)		
		Robust Mean	n used	Robust Uncert.	Robust StDev		Robust Mean	n used	
038.52	Molybdenum/ICP-MS, Open vessel (ppm)	0.9658	2		0.0223		0.0685	2	
038.99	Molybdenum/Miscellaneous (ppm)	1.4	1						
040.53	Barium/ICP-MS, Microwave (46ppm)	5.833	2		0.0176		0.42895	2	
040.43	Barium/ICP, Microwave (46ppm)	6.97	1				0.06	1	
041.53	Vanadium/ICP-MS, Microwave (46ppm)	0.4734	1				0.0413	1	
042.00	Chloride/Titrimetric (%)	0.2841	7	0.0096	0.0204	7.17	4.82	0.0111	5
042.99	Chloride/Miscellaneous (%)	0.304	4	0.0079	0.0127	4.17	4.77	0.0053	3
101.99	Choline Chloride/Miscellaneous (ppm)	1715	1				50	1	
102.01	Niacin/Microbiological (ppm)	85.2	1				2.2	1	
103.01	Pantothenic Acid/Microbiological (ppm)	11.05	1				1.7	1	
104.00	Riboflavin/Fluorometric (ppm)	8.28	1				0.26	1	
105.01	Thiamine/Fluorometer (ppm)	5.24	1				0.24	1	
106.02	Vitamin A/LC (KU/kg)	5.544	6	0.6246	1.224	22.1		0.9964	6
106.00	Vitamin A/Color (KU/kg)	6.22	1				0.36	1	
106.03	Vitamin A/ANKOM-LC (KU/kg)	9.27	1				5.92	1	
106.99	Vitamin A/Miscellaneous (KU/kg)	4.34	1				0.4	1	
106.01	Vitamin A/UV (KU/kg)	3.78	1				1.78	1	
107.00	Vitamin B12/Microbiological (ppb)	48.65	1				0.9	1	
107.99	Vitamin B12/Miscellaneous (ppb)	35.65	1				2.7	1	

Code	Analyte / Method	¹ Trueness (Lab Value)					² Thompson Horwitz %RSD	³ Precision (range)	
		Robust Mean	n used	Robust Uncert.	Robust StDev	Robust %RSD		Robust Mean	n used
108.02	Vitamin D3/LC (KU/kg)	1.755	3	0.4776	0.6618	37.7			
108.99	Vitamin D3/Miscellaneous (KU/kg)	1.936	1				0.498	1	
109.02	Vitamin E/LC (IU/kg)	50.63	6	5.12	10.03	19.8		3.215	6
109.99	Vitamin E/Miscellaneous (IU/kg)	49.5	2		14.14		4	2	
109.03	Vitamin E/ANKOM-LC (IU/kg)	52.3	1				3.6	1	
112.01	Pyridoxine/LC (ppm)	137.5	1				1	1	
113.01	Folic Acid/Micro (ppm)	0.9025	1				0.057	1	
114.99	Biotin/Miscellaneous (ppm)	0.1845	1				0.033	1	
115.00	Non Protein N (NPN)/Urea + Am, Urease method (%)	0.088	1				0.144	1	
120.00	Alanine/Post-col Ninhydrin Der (%)	0.9729	7	0.0109	0.023	2.37	4.01	0.0113	7
120.05	Alanine/Pre-col AQC Der (%)	0.9543	6	0.0237	0.0465	4.87	4.02	0.011	5
120.99	Alanine/Miscellaneous (%)	0.9125	2		0.0035		0.05	1	
120.02	Alanine/Post-col OPA Der (%)	0.96	1				0.004	1	
121.00	Arginine/Post-col Ninhydrin Der (%)	1.038	7	0.0142	0.03	2.89	3.97	0.017	6
121.05	Arginine/Pre-col AQC Der (%)	1.056	7	0.0369	0.0781	7.4	3.96	0.0531	5
121.99	Arginine/Miscellaneous (%)	0.985	2		0.0424		0.06	2	
121.02	Arginine/Post-col OPA Der (%)	1.004	1				0.025	1	
122.05	Aspartic/Pre-col AQC Der (%)	1.632	7	0.0875	0.1852	11.3	3.71	0.0604	5
122.00	Aspartic/Post-col Ninhydrin Der (%)	1.604	7	0.0189	0.0399	2.49	3.72	0.0182	7

Code	Analyte / Method	¹ Trueness (Lab Value)				² Thompson Horwitz %RSD	³ Precision (range)		
		Robust Mean	n used	Robust Uncert.	Robust StDev		Robust Mean	n used	
122.99	Aspartic/Miscellaneous (%)	1.645	2		0.2616		0.04	1	
122.02	Aspartic/Post-col OPA Der (%)	1.71	1				0.023	1	
124.05	Cysteine/Cystine/PAO Pre-col AQC Der (%)	0.2808	7	0.0163	0.0346	12.3	4.83	0.0095	6
124.00	Cysteine/Cystine/PAO Post-col Ninhydrin (%)	0.3128	6	0.0082	0.016	5.12	4.75	0.0076	3
124.02	Cysteine/Cystine/PAO Post-col OPA Der (%)	0.309	2		0.041		0.028	2	
124.99	Cysteine/Cystine/Miscellaneous (%)	0.29	1						
125.00	Glutamic/Post-col Ninhydrin Der (%)	3.243	7	0.0325	0.0687	2.12	3.34	0.0217	6
125.05	Glutamic/Pre-col AQC Der (%)	3.152	7	0.0809	0.1713	5.43	3.36	0.0578	6
125.99	Glutamic/Miscellaneous (%)	3.185	2		0.0354		0.12	2	
125.02	Glutamic/Post-col OPA Der (%)	3.404	1				0.025	1	
126.00	Glycine/Post-col Ninhydrin Der (%)	0.7753	7	0.0117	0.0247	3.18	4.15	0.0152	5
126.05	Glycine/Pre-col AQC Der (%)	0.8106	7	0.0379	0.0801	9.88	4.12	0.0374	6
126.99	Glycine/Miscellaneous (%)	0.685	2		0.0566		0.01	2	
126.02	Glycine/Post-col OPA Der (%)	0.811	1				0.014	1	
127.00	Histidine/Post-col Ninhydrin Der (%)	0.4617	7	0.0131	0.0276	5.98	4.48	0.0097	5
127.05	Histidine/Pre-col AQC Der (%)	0.4544	7	0.0162	0.0343	7.55	4.49	0.024	5
127.99	Histidine/Miscellaneous (%)	0.4675	2		0.0318		0.03	1	
127.02	Histidine/Post-col OPA Der (%)	0.4625	1				0.003	1	
128.00	Isoleucine/Post-col Ninhydrin Der (%)	0.6958	7	0.0227	0.0481	6.92	4.21	0.0215	6

Code	Analyte / Method	¹ Trueness (Lab Value)					² Thompson Horwitz %RSD	³ Precision (range)	
		Robust Mean	n used	Robust Uncert.	Robust StDev	Robust %RSD		Robust Mean	n used
128.05	Isoleucine/Pre-col AQC Der (%)	0.6926	7	0.0129	0.0273	3.94	4.22	0.0333	5
128.99	Isoleucine/Miscellaneous (%)	0.66	2		0.0071			0.01	2
128.02	Isoleucine/Post-col OPA Der (%)	0.7015	1					0.009	1
129.05	Leucine/Pre-col AQC Der (%)	1.525	7	0.0301	0.0637	4.18	3.75	0.0419	5
129.00	Leucine/Post-col Ninhydrin Der (%)	1.514	7	0.0253	0.0536	3.54	3.75	0.0204	6
129.99	Leucine/Miscellaneous (%)	1.565	2		0.0919			0.04	1
129.02	Leucine/Post-col OPA Der (%)	1.478	1					0.001	1
130.00	L-Lysine/Post-col Ninhydrin Der (%)	0.8375	10	0.0168	0.0426	5.08	4.1	0.0125	9
130.05	L-Lysine/Pre-col AQC Der (%)	0.8148	8	0.0321	0.0727	8.92	4.11	0.0407	6
130.99	L-Lysine/Miscellaneous (%)	0.8625	2		0.053			0.03	1
130.02	L-Lysine/Post-col OPA Der (%)	0.8825	1					0.003	1
131.00	Methionine/PAO Post-col Ninhydrin Der (%)	0.311	9	0.0043	0.0104	3.33	4.75	0.0106	8
131.05	Methionine/PAO Pre-col AQC Der (%)	0.2901	7	0.0104	0.022	7.58	4.8	0.0105	6
131.02	Methionine/PAO Post-col OPA Der (%)	0.318	2		0.017			0.02	1
131.99	Methionine/Miscellaneous (%)	0.305	2		0.0424			0.02	2
132.00	Phenylalanine/Post-col Ninhydrin Der (%)	0.834	7	0.0145	0.0306	3.67	4.1	0.017	7
132.05	Phenylalanine/Pre-col AQC Der (%)	0.8072	6	0.0152	0.0297	3.68	4.12	0.0195	5
132.99	Phenylalanine/Miscellaneous (%)	0.79	2		0.0424			0.04	1
132.02	Phenylalanine/Post-col OPA Der (%)	0.822	1					0.004	1

Code	Analyte / Method	¹ Trueness (Lab Value)					² Thompson Horwitz %RSD	³ Precision (range)	
		Robust Mean	n used	Robust Uncert.	Robust StDev	Robust %RSD		Robust Mean	n used
133.00	Proline/Post-col Ninhydrin Der (%)	1.15	7	0.0114	0.0241	2.1	3.91	0.0159	7
133.05	Proline/Pre-col AQC Der (%)	1.171	7	0.0226	0.0479	4.09	3.9	0.0056	5
133.99	Proline/Miscellaneous (%)	1.13	2		0.0849			0.04	1
134.05	Serine/Pre-col AQC Der (%)	0.8843	7	0.0302	0.064	7.23	4.06	0.0248	6
134.00	Serine/Post-col Ninhydrin Der (%)	0.8467	7	0.0132	0.028	3.3	4.09	0.0125	6
134.99	Serine/Miscellaneous (%)	0.835	2		0.0141			0.02	2
134.02	Serine/Post-col OPA Der (%)	0.736	1					0.002	1
135.00	Threonine/Post-col Ninhydrin Der (%)	0.6651	7	0.0101	0.0213	3.21	4.24	0.0102	6
135.05	Threonine/Pre-col AQC Der (%)	0.6806	7	0.023	0.0486	7.14	4.23	0.0334	5
135.99	Threonine/Miscellaneous (%)	0.665	2		0.0141			0.02	2
135.02	Threonine/Post-col OPA Der (%)	0.6415	1					0.003	1
136.05	Tryptophan/Pre-col AQC Der (%)	0.1855	3	0.0274	0.038	20.5	5.14	0.0072	3
136.03	Tryptophan/Alka-Hydrol + IS RP LC FI (%)	0.2216	3	0.0166	0.023	10.4	5		
136.99	Tryptophan/Miscellaneous (%)	0.3445	3	0.166	0.23	66.8	4.68		
136.01	Tryptophan/Alka-Hydrol Rev Phase LC UV (%)	0.2275	2		0.0035			0.01	1
136.02	Tryptophan/Alka-Hydrol Post-col OPA De (%)	0.1901	1					0.0002	1
137.05	Tyrosine/Pre-col AQC Der (%)	0.5839	7	0.0334	0.0707	12.1	4.33	0.0458	5
137.00	Tyrosine/Post-col Ninhydrin Der (w/o oxidation) (%)	0.553	6	0.0431	0.0845	15.3	4.36	0.0092	6
137.99	Tyrosine/Miscellaneous (%)	0.5075	2		0.0742			0.025	2

Code	Analyte / Method	¹ Trueness (Lab Value)					² Thompson Horwitz %RSD	³ Precision (range)	
		Robust Mean	n used	Robust Uncert.	Robust StDev	Robust %RSD		Robust Mean	n used
137.02	Tyrosine/Post-col OPA Der (%)	0.4635	1					0.037	1
138.00	Valine/Post-col Ninhydrin Der (%)	0.8313	7	0.0259	0.0548	6.59	4.1	0.0199	7
138.05	Valine/Pre-col AQC Der (%)	0.8156	7	0.0159	0.0336	4.12	4.11	0.0308	5
138.99	Valine/Miscellaneous (%)	0.7825	2		0.0035			0.015	2
138.02	Valine/Post-col OPA Der (%)	0.906	1					0.008	1
139.00	Taurine/Post-col Ninhydrin Der (%)	0.15	1						
139.05	Taurine/Pre-col AQC Der (%)	0.0285	1					0.001	1
160.10	Fructose/HPAEC PAD (%)	0.2425	2		0.0884			0.015	2
160.99	Fructose/Miscellaneous (%)	0.315	1					0.03	1
162.10	Glucose/HPAEC PAD (%)	0.2675	2		0.0672			0.03	1
162.99	Glucose/Miscellaneous (%)	0.305	1					0.05	1
163.10	Lactose/HPAEC PAD (%)	0.09	1						
164.10	Maltose/HPAEC PAD (%)	0.3225	2		0.1732			0.05	1
165.10	Sucrose/HPAEC PAD (%)	1.92	2		0.099			0.1	2
165.99	Sucrose/Miscellaneous (%)	2.065	1					0.05	1
166.10	Raffinose/HPAEC PAD (%)	0.4	2		0			0.04	2
166.99	Raffinose/Miscellaneous (%)	0.345	1					0.01	1
167.10	Stachyose/HPAEC PAD (%)	1.1	2		0.0283			0.04	2
167.99	Stachyose/Miscellaneous (%)	1.8	1					0.02	1

Code	Analyte / Method	¹ Trueness (Lab Value)				² Thompson Horwitz %RSD	³ Precision (range)		
		Robust Mean	n used	Robust Uncert.	Robust StDev		Robust Mean	n used	
345.02	Amprolium/LC (UV or FL) (ppm)	108.9	7	5.043	10.67	9.81	7.86	3.814	6
345.00	Amprolium/Colorimetric (ppm)	127.7	4	11.94	19.1	15	7.68	1.867	3
400.01	Water Activity/Aqualab chilled mirror (Units)	0.6407	6	0.007	0.0137	2.13	1.25	0.0036	6
400.99	Water Activity/Miscellaneous (Units)	0.7	1						
516.53	Arsenic, Total/ICP-MS, Microwave (ppm)	0.0716	3	0.0307	0.0425	59.4	22	0.0018	3
516.00	Arsenic, Total/AA, Hydride (ppm)	0.0525	1					0.009	1
518.53	Cadmium/ICP-MS, Microwave (ppm)	0.0578	3	0.0082	0.0113	19.6	22	0.0043	3
518.43	Cadmium/ICP, Microwave (ppm)	0.1153	2		0.0208			0.04195	2
518.41	Cadmium/ICP, Dry ash (ppm)	0.061	1					0.0056	1
518.52	Cadmium/ICP-MS, Open vessel (ppm)	0.0545	1					0.0014	1
518.33	Cadmium/AAS, Microwave (ppm)	0.0542	1					0.0072	1
520.43	Chromium/ICP, Microwave (ppm)	2.288	5	0.1905	0.3408	14.9	14	0.036	4
520.53	Chromium/ICP-MS, Microwave (ppm)	2.194	3	0.4618	0.6399	29.2	14.1	0.2804	3
520.52	Chromium/ICP-MS, Open vessel (ppm)	1.715	2		0.3041			0.06	2
520.42	Chromium/ICP, Open vessel (ppm)	2.148	2		0.1308			0.225	2
520.41	Chromium/ICP, Dry ash (ppm)	1.716	1					0.3654	1
526.53	Lead/ICP-MS, Microwave (ppm)	0.0595	3	0.0042	0.0058	9.78	22	0.004	3
526.43	Lead/ICP, Microwave (ppm)	1.435	2		1.684			1.17745	2
526.33	Lead/AAS, Microwave (ppm)	0.4698	1					0.0472	1

Code	Analyte / Method	¹ Trueness (Lab Value)				² Thompson Horwitz %RSD	³ Precision (range)		
		Robust Mean	n used	Robust Uncert.	Robust StDev		Robust Mean	n used	
526.52	Lead/ICP-MS, Open vessel (ppm)	0.0636	1				0.0006	1	
526.41	Lead/ICP, Dry ash (ppm)	0.1594	1				0.0582	1	
529.99	Mercury/Miscellaneous (ppb)	5.981	1				0.279	1	
539.43	Nickel/ICP, Microwave (ppm)	3.032	4	1.186	1.897	62.6	13.4	0.3722	4
539.53	Nickel/ICP-MS, Microwave (ppm)	2.537	2		0.5909			0.19185	2
539.41	Nickel/ICP, Dry ash (ppm)	1.768	1					0.0797	1
539.52	Nickel/ICP-MS, Open vessel (ppm)	2.29	1					0.08	1
706.99	Caprylic acid (8:0)/Miscellaneous (% w/w)	0.0055	1					0.009	1
708.99	Capric acid (10:0)/Miscellaneous (% w/w)	0.004	1					0.006	1
710.99	Lauric Acid (12:0)/Miscellaneous (% w/w)	0.0035	1					0.003	1
714.99	Myristic Acid (14:0)/Miscellaneous (% w/w)	0.005	1					0.002	1
716.99	Palmitic Acid (16:0)/Miscellaneous (% w/w)	0.4457	2		0.0103			0.02435	2
718.99	Palmitoleic Acid (9c-16:1)/Miscellaneous (% w/w)	0.0055	1					0.001	1
720.99	Margaric acid (17:0)/Miscellaneous (% w/w)	0.0045	1					0.001	1
722.99	Stearic Acid (18:0)/Miscellaneous (% w/w)	0.0744	2		0.0037			0.00525	2
724.99	Oleic Acid (9c-18:1)/Miscellaneous (% w/w)	0.5867	2		0.0258			0.01765	2
726.99	Linoleic Acid (9c,12c-18:2)/Miscellaneous (% w/w)	1.47	2		0.0337			0.0484	2
728.99	alpha-Linolenic Acid (9c,12c,15c-18:3)/Miscellaneous (% w/w)	0.0863	2		0.0004			0.0013	1
730.99	Arachidic Acid (20:0)/Miscellaneous (% w/w)	0.0104	2		0.0002			0.00055	2

Code	Analyte / Method	¹ Trueness (Lab Value)				² Thompson Horwitz %RSD	³ Precision (range)	
		Robust Mean	n used	Robust Uncert.	Robust StDev		Robust Mean	n used
732.99	Gondoic Acid (11c-20:1)/Miscellaneous (% w/w)	0.0086	2		0.0001		0.001	1
742.99	Behenic Acid (22:0)/Miscellaneous (% w/w)	0.009	2		0.0014		0.0013	2
744.99	Erucic Acid (13c-22:1)/Miscellaneous (% w/w)	0.002	1					
748.99	Lignoceric Acid (24:0)/Miscellaneous (% w/w)	0.0106	2		0.0005		0.00105	2
750.99	Docosahexaenoic Acid DHA (4c,7c,10c,13c,16c,19c-22/Miscellaneous	0.0015	1				0.001	1
752.99	Nervonic Acid (24:1) isomers/Miscellaneous (% w/w)	0.0035	1				0.001	1
754.99	Total n-3 Polyunsaturated (Omega-3) Fatty Acids/Miscellaneous (%)	0.088	1				0.002	1
756.99	Total n-6 Polyunsaturated (Omega-6) Fatty Acids/Miscellaneous (%)	1.454	1				0.099	1
758.99	Total Saturated Fatty Acids/Miscellaneous (% w/w)	0.5865	1				0.067	1
764.99	Total cis Monounsaturated Fatty Acids/Miscellaneous (% w/w)	0.6235	1				0.029	1
768.99	Total cis Polyunsaturated Fatty Acids/Miscellaneous (% w/w)	1.546	1				0.095	1
770.99	Total Fat (equivalent to NLEA)/Miscellaneous (% w/w)	2.893	1				0.2	1
772.99	Total Fatty Acids/Miscellaneous (% w/w)	2.75	2		0.0216		0.1105	2

1. Trueness Parameters: Statistical parameters defining the distribution of lab values which are used to evaluate how close a Lab Value is to the mean. Parameters are shown for number of observations used (n used). Method All Labs PT report identifies data not used. Robust statistics was employed to determine mean and standard deviation (StDev) if number of observations used (n used) ≥ 6 (blue background). Classical statistics was employed if number of observations used (n used) < 6 (no color background). %RSD is the relative standard deviation with respect to the mean ($\text{StDev}/\text{Mean} \times 100$). Uncertainty (Uncert.) is a measure of where the true population mean lies.

2. Thompson-Horwitz %RSD: Expected relative standard deviation based on analysis of data by Thompson and Horwitz (Thompson, DOI: 10.1039/b000282h).

3. Precision Parameters: Lab's precision is estimated by the difference in 2 results reported by a lab (range). Mean of ranges are shown for number of observations used (n used). Method All Tests report identifies data not used. Robust statistics was employed to determine mean if number of observations used (n used) ≥ 6 (green background). Classical statistics was employed if number of observations used (n used) < 6 (no color background).

Appendix

Content Description of METHOD Summary Statistics Report

The Method Summary Statistics Report provides trueness and precision parameters from determination of analytes by specific methods. Determination of summary statistics followed protocols in ISO 13528:2015(E) using Algorithm A robust analysis (Statistical methods for use in proficiency testing by interlaboratory comparison). Robust statistics was used to determine statistical parameters for sets with 6 or more observations. Classical statistics was used for sets with 3, 4, or 5 observations. Robust statistics has an advantage of removing undesired influence of outlying data on the mean and standard deviation without removing data from the statistical analysis.

For trueness, the mean and standard deviation are presented for the number of observations (n used). The uncertainty (Uncert.) is a measure of where the “real” value for the concentration lies above or below the mean with a 68% certainty ($1.25 * \text{standard deviation} / (n \text{ used})^{0.5}$). As the number of observations (n used) increases, uncertainty decreases. The relative standard deviation (%RSD) is a percentage of the standard deviation divided by the mean ($\text{standard deviation} / \text{mean} \times 100$). The Thompson-Horwitz %RSD is a standard benchmark on variability based on data analyzed by Thompson and Horwitz (Thompson, DOI: 10.1039/b000282h).

Precision in the data populations is estimated by the range of duplicate results reported. The robust or classical mean is presented along with the number of observations. Any duplicate results that are exactly the same are removed in the determination of the mean to remove undue influence of entries that may be from labs reporting one result twice.